3295	1579.4	124.52	542.72
#1	.15485	.00380	.00035	25.792	-.00140	.01160	.00004
#2	.04609	.00044	.00064	25.741	-.00007	.02911	-.00462
#3	-.02785	.00489	.00043	26.363	.00122	-.00253	.00817
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606077911SDL Acquired: 6/23/2016 19:54:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573973-02

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00030**	**-.00465**	**.06796**	**.00020**	**.00397**	**-.00466**	**-.00173**
Stddev	.00330	.00198	.00111	.00072	.00038	.00590	.00388
%RSD	1090.6	42.606	1.6393	364.61	9.6242	126.80	224.33
#1	.00173	-.00263	.06819	.00039	.00374	-.01146	-.00420
#2	.00148	-.00660	.06894	-.00060	.00441	-.00094	-.00374
#3	-.00411	-.00472	.06675	.00080	.00377	-.00157	.00274
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.00000**	**.00179**	F **-.14942**
Stddev	.00024	.00022	.27065
%RSD	17591.	12.069	181.14
#1	-.00010	.00175	-.05629
#2	-.00017	.00203	-.45434
#3	.00028	.00160	.06237
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6956.7**	**42908.**	**2182.4**
Stddev	24.4	138.	8.1
%RSD	.35082	.32162	.37002
#1	6945.3	43022.	2184.2
#2	6984.7	42948.	2173.6
#3	6940.1	42755.	2189.5

Sample Name: L1606077911SDL Acquired: 6/23/2016 19:58:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573973-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00091**	**.00551**	**-.00163**	**.00298**	**.00016**	**-.00004**	**.14622**	**-.00004**
Stddev	.00035	.00230	.00441	.00295	.00019	.00009	.00718	.00022
%RSD	37.880	41.654	270.73	98.999	121.74	255.43	4.9086	633.58
#1	-.00082	.00812	-.00328	.00338	.00020	.00005	.14190	-.00029
#2	-.00130	.00463	.00337	-.00015	.00033	-.00003	.14227	.00014
#3	-.00062	.00379	-.00498	.00571	-.00005	-.00013	.15451	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**-.00012**	**.00129**	**.01297**	**.07882**	**-.00139**	**.02445**	**.00297**
Stddev	.00029	.00066	.00139	.04085	.05847	.00356	.09577	.00349
%RSD	220.77	550.63	108.07	315.10	74.183	257.22	391.63	117.45
#1	-.00041	-.00064	.00208	-.02649	.01827	-.00070	-.02314	-.00101
#2	.00018	-.00034	.00210	.01030	.08323	.00179	-.03820	.00442
#3	-.00016	.00062	-.00032	.05509	.13495	-.00524	.13470	.00550
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00006**	**4.6601**	**.00016**	**.01099**	**-.00374**	**-.00006**	**-.00666**	**-.00121**
Stddev	.00064	.0487	.00101	.00932	.00214	.00271	.00377	.00406
%RSD	1145.3	1.0448	617.86	84.814	57.083	4166.1	56.580	334.61
#1	-.00077	4.6317	.00117	.02130	-.00354	.00054	-.01063	.00307
#2	.00017	4.7163	-.00086	.00316	-.00172	-.00302	-.00622	-.00502
#3	.00043	4.6322	.00019	.00850	-.00597	.00229	-.00313	-.00169
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606077911SDL Acquired: 6/23/2016 19:58:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573973-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00037**	**.00016**	**.00324**	**-.00190**	**.00015**	**.00102**	**.45592**
Stddev	.00090	.00024	.00774	.00281	.00036	.00011	.15548
%RSD	242.24	152.87	238.78	147.46	242.78	10.924	34.103
#1	-.00010	-.00008	-.00569	-.00173	-.00002	.00104	.63162
#2	.00036	.00041	.00759	.00081	.00057	.00089	.33613
#3	-.00138	.00014	.00782	-.00479	-.00009	.00111	.40000
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**7002.1**	**43421.**	**2180.6**
Stddev	12.5	150.	18.1
%RSD	.17813	.34564	.82992
#1	7007.5	43338.	2178.3
#2	6987.9	43594.	2163.7
#3	7011.0	43331.	2199.7

Sample Name: CCV Acquired: 6/23/2016 20:02:38 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.38352	9.5676	.38149	.48360	.95868	.04827	9.5943
Stddev	.00322	.0401	.00429	.00105	.00492	.00022	.0294
%RSD	.83982	.41946	1.1243	.21691	.51367	.46090	.30624
#1	.38657	9.6123	.37703	.48417	.95340	.04832	9.5754
#2	.38382	9.5347	.38559	.48239	.95948	.04803	9.6282
#3	.38015	9.5558	.38184	.48424	.96315	.04846	9.5794
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04762	.19248	.48389	.48328	3.8584	48.266	.97868
Stddev	.00015	.00028	.00059	.00212	.0428	.089	.00720
%RSD	.31413	.14541	.12117	.43944	1.1080	.18417	.73546
#1	.04779	.19230	.48370	.48346	3.8177	48.266	.97363
#2	.04758	.19280	.48342	.48107	3.8546	48.177	.97549
#3	.04750	.19234	.48455	.48530	3.9029	48.355	.98692
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.6513	.48380	.96301	47.927	.48135	9.5102	.48331
Stddev	.2636	.00656	.00332	.072	.00114	.0199	.00231
%RSD	2.7313	1.3558	.34426	.15061	.23638	.20964	.47887
#1	9.3550	.47806	.96658	47.845	.48232	9.5216	.48282
#2	9.8597	.48238	.96243	47.981	.48165	9.4872	.48129
#3	9.7393	.49095	.96003	47.956	.48010	9.5218	.48584
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 20:02:38 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1479**	**.38358**	**4.8260**	**.95697**	**.96108**	**.96615**	**.48705**
Stddev	.0066	.01108	.0032	.00258	.00341	.00961	.00179
%RSD	.57598	2.8894	.06586	.26977	.35526	.99472	.36795
#1	1.1549	.37665	4.8269	.95791	.95718	.95798	.48502
#2	1.1417	.39636	4.8286	.95405	.96254	.96374	.48839
#3	1.1472	.37773	4.8224	.95895	.96352	.97674	.48775
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.96626**	**.96309**	F **.37739**
Stddev	.00265	.00068	.17080
%RSD	.27390	.07033	45.257
#1	.96486	.96244	.28567
#2	.96461	.96302	.27206
#3	.96931	.96379	.57445
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6922.0**	**42512.**	**2201.6**
Stddev	21.1	353.	11.3
%RSD	.30486	.83050	.51529
#1	6934.6	42346.	2213.8
#2	6933.8	42273.	2199.8
#3	6897.7	42918.	2191.3

Sample Name: CCB Acquired: 6/23/2016 20:06:16 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00019	-.00041	.00090	.00065	-.00017	-.00009	.03364	-.00010
Stddev	.00250	.00080	.00202	.00197	.00087	.00003	.02458	.00030
%RSD	1325.3	193.76	223.31	305.36	513.01	38.963	73.061	282.77
#1	-.00159	-.00083	-.00002	.00192	-.00117	-.00006	.00559	-.00032
#2	.00304	.00051	-.00049	.00164	.00038	-.00008	.04390	.00023
#3	-.00089	-.00092	.00322	-.00163	.00029	-.00012	.05142	-.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00011	.00089	.00107	.02622	.14403	.00140	.16533	.00084
Stddev	.00031	.00098	.00055	.02571	.09143	.00408	.10839	.00123
%RSD	276.41	109.41	51.427	98.056	63.481	291.75	65.560	147.25
#1	.00047	.00000	.00094	-.00236	.13572	.00290	.28604	.00006
#2	-.00006	.00194	.00060	.03356	.05704	.00451	.07632	.00226
#3	-.00008	.00074	.00168	.04747	.23934	-.00322	.13363	.00020
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00296	.11509	-.00013	.00085	.00094	.00629	-.00145	.00187
Stddev	.00021	.03731	.00042	.00219	.00279	.00604	.00094	.00124
%RSD	7.1713	32.413	339.72	257.22	298.65	96.053	64.718	66.434
#1	.00285	.13809	-.00049	.00212	.00095	.00557	-.00045	.00293
#2	.00321	.07205	.00034	.00212	.00372	.01265	-.00231	.00218
#3	.00283	.13514	-.00022	-.00168	-.00186	.00064	-.00160	.00050
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: CCB Acquired: 6/23/2016 20:06:16 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00007**	**-.00022**	**-.00145**	**-.00171**	**-.00068**	**.00008**	**.01839**
Stddev	.00140	.00022	.01047	.00376	.00047	.00019	.18176
%RSD	1871.4	101.49	723.55	219.98	69.299	243.52	988.14
#1	-.00163	-.00007	.01060	.00231	-.00016	.00013	.02617
#2	.00109	-.00047	-.00838	-.00514	-.00082	.00023	.19614
#3	.00032	-.00011	-.00656	-.00229	-.00108	-.00013	-.16713
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6970.0**	**43262.**	**2192.2**
Stddev	74.7	174.	9.5
%RSD	1.0710	.40250	.43124
#1	7003.8	43124.	2200.3
#2	7021.8	43458.	2194.3
#3	6884.5	43203.	2181.8

Sample Name: L1606077914 Acquired: 6/23/2016 20:10:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00149	.10628	.00096	.02840	.05288	-.00005	7.4368	.00027
Stddev	.00097	.00356	.00214	.00184	.00085	.00008	.1298	.00005
%RSD	64.971	3.3536	223.67	6.4703	1.6095	158.12	1.7453	17.347
#1	.00054	.10724	-.00146	.02628	.05351	-.00006	7.5825	.00032
#2	.00248	.10234	.00174	.02939	.05191	-.00012	7.3941	.00025
#3	.00145	.10927	.00259	.02952	.05322	.00003	7.3337	.00024
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00039	.00052	.00185	.08486	.70361	-.00341	1.3249	.01531
Stddev	.00024	.00198	.00037	.02791	.04945	.00257	.0384	.00314
%RSD	62.961	378.91	20.059	32.893	7.0277	75.397	2.8943	20.533
#1	.00050	.00276	.00177	.06380	.69565	-.00073	1.3540	.01542
#2	.00055	-.00104	.00226	.11652	.75656	-.00364	1.3392	.01211
#3	.00011	-.00015	.00153	.07425	.65863	-.00587	1.2814	.01840
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	136.97	.00087	.00420	.00004	.00045	-.00208	.33477
Stddev	.00023	.63	.00055	.00550	.00261	.00047	.00600	.00123
%RSD	39.437	.45874	63.261	131.00	7265.2	105.14	288.25	.36659
#1	.00032	137.68	.00147	.00987	-.00284	.00003	.00096	.33486
#2	.00071	136.77	.00073	.00385	.00070	.00095	.00179	.33594
#3	.00074	136.47	.00040	-.00112	.00225	.00036	-.00900	.33350
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606077914 Acquired: 6/23/2016 20:10:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00035	.03351	-.00005	-.00700	-.00055	.00972	.13105
Stddev	.00062	.00057	.01222	.00229	.00057	.00011	.18984
%RSD	177.30	1.6905	24453.	32.742	102.06	1.1714	144.86
#1	-.00033	.03393	.00766	-.00548	.00007	.00981	.20911
#2	.00048	.03373	-.01414	-.00589	-.00071	.00959	.26943
#3	.00090	.03286	.00633	-.00964	-.00102	.00975	-.08538
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6756.2	41698.	2168.9
Stddev	15.0	223.	10.5
%RSD	.22178	.53392	.48265
#1	6768.0	41441.	2173.7
#2	6739.3	41816.	2176.1
#3	6761.2	41837.	2156.9

Sample Name: L1606080801 Acquired: 6/23/2016 20:14:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.00981	.00166	.02700	.00536	-.00004	53.221	.00023
Stddev	.00132	.00387	.00350	.00355	.00074	.00005	.458	.00031
%RSD	223.52	39.428	210.37	13.162	13.862	133.64	.86119	137.26
#1	-.00079	.01246	.00412	.02618	.00458	.00002	53.750	-.00010
#2	.00184	.00537	-.00234	.03090	.00606	-.00008	52.944	.00025
#3	.00072	.01160	.00322	.02393	.00543	-.00006	52.968	.00052
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00007	-.00019	.00217	.26849	1.4108	-.00252	15.011	.12291
Stddev	.00031	.00023	.00068	.03341	.0586	.00219	.184	.00135
%RSD	424.72	122.66	31.151	12.444	4.1556	86.964	1.2259	1.0966
#1	-.00007	-.00003	.00235	.24539	1.3436	-.00503	15.179	.12150
#2	-.00038	-.00008	.00142	.25327	1.4510	-.00152	15.040	.12419
#3	.00023	-.00045	.00273	.30679	1.4380	-.00101	14.814	.12302
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00051	15.947	.00061	.00603	.00298	.00330	-.00384	5.1133
Stddev	.00030	.171	.00079	.00285	.00319	.00406	.01618	.0046
%RSD	57.821	1.0703	128.91	47.356	107.05	123.06	420.92	.09015
#1	.00017	16.102	-.00004	.00833	.00524	-.00067	-.02178	5.1184
#2	.00071	15.975	.00148	.00284	.00438	.00312	.00059	5.1096
#3	.00066	15.764	.00039	.00691	-.00067	.00744	.00965	5.1119
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606080801 Acquired: 6/23/2016 20:14:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00027**	**.17284**	**-.01553**	**-.00411**	**-.00044**	**.00368**	**.21621**
Stddev	.00054	.00134	.00554	.00196	.00032	.00035	.15335
%RSD	203.08	.77573	35.693	47.631	72.739	9.4159	70.926
#1	-.00035	.17423	-.01858	-.00259	-.00007	.00340	.36753
#2	.00031	.17155	-.00913	-.00632	-.00067	.00406	.22018
#3	-.00077	.17274	-.01887	-.00342	-.00059	.00356	.06091
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6663.3**	**41709.**	**2149.6**
Stddev	10.8	204.	16.5
%RSD	.16157	.48945	.76736
#1	6671.1	41908.	2131.3
#2	6667.7	41719.	2154.5
#3	6651.0	41500.	2163.2

Sample Name: L1606080803S Acquired: 6/23/2016 20:18:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20531	4.9958	.20276	1.0092	.51435	.02483	57.046	.02496
Stddev	.00292	.0043	.00269	.0024	.00665	.00006	.768	.00030
%RSD	1.4202	.08567	1.3273	.24307	1.2934	.23022	1.3467	1.1942
#1	.20859	4.9909	.20264	1.0074	.52180	.02478	57.925	.02528
#2	.20436	4.9977	.20014	1.0120	.50900	.02489	56.503	.02470
#3	.20300	4.9987	.20552	1.0083	.51225	.02482	56.709	.02490
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09967	.25096	.25022	2.3265	26.678	.52152	20.005	.37815
Stddev	.00055	.00036	.00033	.0587	.264	.00877	.198	.00476
%RSD	.55302	.14274	.13074	2.5243	.98845	1.6818	.98772	1.2593
#1	.10016	.25128	.25042	2.3809	26.980	.53116	20.035	.38218
#2	.09976	.25104	.24984	2.2643	26.497	.51402	19.794	.37290
#3	.09907	.25057	.25040	2.3343	26.556	.51939	20.186	.37938
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.52098	41.169	.25038	4.9597	.24975	.60811	.19088	7.6345
Stddev	.00121	.394	.00120	.0047	.00531	.00285	.00510	.0102
%RSD	.23295	.95680	.47973	.09548	2.1271	.46887	2.6703	.13352
#1	.52204	41.621	.25177	4.9644	.24365	.60722	.18517	7.6439
#2	.51965	40.900	.24967	4.9597	.25336	.61129	.19249	7.6237
#3	.52125	40.985	.24970	4.9549	.25223	.60580	.19497	7.6359
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606080803S Acquired: 6/23/2016 20:18:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-07

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50663	.69353	.50821	.24979	.50502	.50112	.56844
Stddev	.00206	.00495	.00984	.00132	.00161	.00059	.24983
%RSD	.40682	.71328	1.9372	.52651	.31873	.11783	43.949
#1	.50895	.69918	.50790	.24829	.50330	.50110	.76022
#2	.50591	.69001	.49852	.25075	.50528	.50054	.28593
#3	.50502	.69139	.51821	.25034	.50649	.50172	.65917
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6625.1	41184.	2141.2
Stddev	43.1	220.	34.4
%RSD	.65093	.53421	1.6047
#1	6575.6	40948.	2101.5
#2	6654.8	41222.	2162.6
#3	6644.9	41383.	2159.3

Sample Name: L1606080804SD Acquired: 6/23/2016 20:21:47 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20417	5.0253	.20216	1.0213	.52389	.02512	58.507	.02515
Stddev	.00061	.0013	.00476	.0019	.00456	.00016	.379	.00007
%RSD	.29699	.02608	2.3559	.18158	.87021	.62066	.64724	.26301
#1	.20348	5.0268	.19939	1.0192	.52806	.02496	58.865	.02518
#2	.20461	5.0245	.20766	1.0224	.52461	.02527	58.546	.02507
#3	.20443	5.0246	.19943	1.0224	.51902	.02514	58.111	.02519
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10013	.25277	.25192	2.3449	27.111	.52740	20.302	.37943
Stddev	.00006	.00076	.00111	.0178	.231	.00552	.216	.00380
%RSD	.06407	.29922	.44124	.75833	.85294	1.0468	1.0620	1.0007
#1	.10006	.25267	.25295	2.3582	27.342	.53203	20.543	.38169
#2	.10018	.25357	.25074	2.3247	27.111	.52888	20.233	.38156
#3	.10015	.25207	.25208	2.3517	26.880	.52129	20.129	.37505
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.52376	41.862	.25037	4.9891	.25282	.61284	.19379	7.7402
Stddev	.00119	.339	.00131	.0038	.00458	.00414	.01024	.0102
%RSD	.22762	.81017	.52333	.07550	1.8112	.67508	5.2847	.13142
#1	.52513	42.192	.25099	4.9856	.25772	.61266	.19670	7.7511
#2	.52310	41.879	.25125	4.9931	.24866	.60879	.20226	7.7310
#3	.52304	41.514	.24886	4.9887	.25208	.61705	.18241	7.7384
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606080804SD Acquired: 6/23/2016 20:21:47 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-08

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50881	.70095	.50511	.25346	.50822	.50328	.36568
Stddev	.00118	.00336	.00758	.00055	.00107	.00052	.09051
%RSD	.23262	.47943	1.5010	.21780	.21111	.10273	24.750
#1	.50993	.70351	.51202	.25388	.50727	.50277	.45775
#2	.50891	.70221	.50631	.25366	.50938	.50380	.36245
#3	.50757	.69715	.49700	.25283	.50801	.50326	.27683
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6648.3	41256.	2167.1
Stddev	37.5	212.	35.6
%RSD	.56386	.51441	1.6437
#1	6687.1	41426.	2128.5
#2	6645.3	41018.	2174.1
#3	6612.3	41324.	2198.7

Sample Name: L1606082202 Acquired: 6/23/2016 20:25:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00152	.02026	-.00003	.10248	.15200	.00003	6.5973	.00005
Stddev	.00174	.00884	.00264	.00097	.00250	.00003	.1001	.00016
%RSD	114.35	43.656	8042.8	.94662	1.6469	112.93	1.5178	320.68
#1	-.00041	.02197	.00037	.10136	.15374	.00007	6.6625	.00019
#2	.00200	.02813	-.00285	.10296	.15311	.00001	6.6474	.00008
#3	.00298	.01069	.00238	.10312	.14913	.00001	6.4820	-.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	-.00052	.00142	2.2172	1.7251	.02347	4.6721	.08008
Stddev	.00019	.00160	.00145	.0203	.1370	.00447	.1537	.00373
%RSD	39.028	307.12	102.30	.91315	7.9439	19.041	3.2904	4.6534
#1	.00065	-.00214	.00215	2.1938	1.8809	.02224	4.7545	.08412
#2	.00052	.00105	-.00025	2.2287	1.6712	.01975	4.7670	.07933
#3	.00028	-.00047	.00236	2.2290	1.6232	.02843	4.4947	.07679
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00084	87.069	.00079	.09610	.00066	.00367	-.00338	12.803
Stddev	.00076	.850	.00109	.00417	.00060	.00400	.00476	.010
%RSD	90.058	.97655	138.31	4.3415	91.329	109.21	141.00	.07526
#1	-.00000	87.844	-.00046	.09129	-.00003	.00288	-.00767	12.795
#2	.00146	87.204	.00156	.09839	.00108	.00011	.00175	12.801
#3	.00107	86.159	.00126	.09864	.00092	.00800	-.00421	12.814
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606082202 Acquired: 6/23/2016 20:25:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00121	.22077	-.00522	-.00363	.00085	.00236	.22639
Stddev	.00050	.00200	.00719	.00112	.00069	.00021	.20548
%RSD	41.542	.90585	137.85	30.833	80.735	8.9617	90.766
#1	-.00156	.22257	-.00824	-.00440	.00009	.00222	.33820
#2	-.00063	.22113	-.01041	-.00235	.00143	.00261	.35171
#3	-.00143	.21862	.00299	-.00415	.00104	.00226	-.01076
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6664.8	41248.	2126.9
Stddev	26.9	323.	26.2
%RSD	.40377	.78403	1.2328
#1	6695.8	40876.	2101.3
#2	6647.4	41460.	2125.7
#3	6651.2	41408.	2153.7

Sample Name: L1606082203 Acquired: 6/23/2016 20:29:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00153	.05110	.00192	.03595	.79888	.00004	32.443	.00019
Stddev	.00257	.00394	.00088	.00096	.00490	.00006	.223	.00012
%RSD	168.19	7.7097	45.736	2.6764	.61347	156.15	.68852	63.266
#1	.00450	.04867	.00097	.03585	.80394	.00006	32.700	.00006
#2	.00008	.05565	.00210	.03696	.79416	-.00003	32.293	.00030
#3	.00001	.04899	.00269	.03505	.79852	.00008	32.336	.00020
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00307	.00707	.00126	5.3792	51.825	.10776	15.276	.18167
Stddev	.00032	.00094	.00080	.0419	.270	.00194	.142	.00297
%RSD	10.345	13.245	64.108	.77858	.52117	1.8035	.92803	1.6345
#1	.00299	.00617	.00218	5.4159	52.078	.10909	15.426	.17840
#2	.00280	.00804	.00087	5.3881	51.541	.10866	15.144	.18240
#3	.00342	.00701	.00071	5.3336	51.857	.10553	15.259	.18420
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00008	93.634	.00315	-.00252	.00643	.00340	-.00637	11.872
Stddev	.00025	.401	.00140	.00666	.00138	.00643	.00382	.014
%RSD	298.72	.42875	44.327	264.34	21.428	189.23	59.894	.11647
#1	-.00020	94.085	.00392	-.00701	.00774	.00248	-.00439	11.857
#2	.00021	93.316	.00154	.00513	.00656	.01023	-.00396	11.885
#3	.00024	93.502	.00400	-.00568	.00499	-.00252	-.01077	11.874
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606082203 Acquired: 6/23/2016 20:29:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00017**	**.82666**	**-.00447**	**-.00598**	**-.00056**	**.00596**	**.27009**
Stddev	.00125	.01637	.00340	.00085	.00033	.00013	.03306
%RSD	751.65	1.9807	76.005	14.225	59.976	2.2544	12.242
#1	.00127	.84555	-.00552	-.00599	-.00043	.00586	.28137
#2	-.00101	.81782	-.00722	-.00682	-.00030	.00611	.29603
#3	-.00076	.81660	-.00067	-.00512	-.00093	.00591	.23286
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6576.7**	**40904.**	**2132.1**
Stddev	24.6	214.	16.8
%RSD	.37432	.52260	.78563
#1	6552.2	40790.	2112.8
#2	6576.5	40771.	2141.0
#3	6601.4	41150.	2142.6

Sample Name: L1606082302 Acquired: 6/23/2016 20:33:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00184	.02080	.00098	.02231	7.8783	.00017	255.75
Stddev	.00203	.01187	.00435	.00335	.0432	.00007	1.49
%RSD	110.13	57.087	445.14	15.024	.54831	41.236	.58087
#1	-.00150	.00710	.00584	.02555	7.9220	.00012	256.99
#2	-.00401	.02815	-.00256	.01885	7.8356	.00014	254.11
#3	-.00000	.02715	-.00034	.02253	7.8772	.00025	256.16
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00094	.05342	.00242	.00486	4.7469	3.4904	.13833
Stddev	.00052	.00070	.00059	.00069	.0046	.0030	.00615
%RSD	55.632	1.3175	24.430	14.175	.09644	.08679	4.4491
#1	.00040	.05380	.00272	.00415	4.7522	3.4870	.14015
#2	.00098	.05261	.00174	.00553	4.7443	3.4928	.13147
#3	.00144	.05385	.00281	.00490	4.7442	3.4915	.14336
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	198.54	.84118	-.00019	F 709.80	.01731	.04470	.00329
Stddev	1.62	.00623	.00016	4.09	.00125	.01000	.00583
%RSD	.81609	.74094	85.381	.57589	7.2164	22.367	177.43
#1	200.41	.84837	-.00014	712.90	.01830	.03650	.00825
#2	197.48	.83796	-.00006	705.16	.01772	.04176	.00474
#3	197.74	.83722	-.00036	711.32	.01590	.05584	-.00313
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Sample Name: L1606082302 Acquired: 6/23/2016 20:33:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00699**	**-.00211**	**18.432**	**-.00053**	**5.9869**	F **-.04100**	**-.00343**
Stddev	.00326	.00362	.021	.00131	.0347	.00868	.00076
%RSD	46.731	171.57	.11640	247.46	.57940	21.163	22.234
#1	.00892	.00180	18.414	-.00131	6.0214	-.04734	-.00288
#2	.00882	-.00536	18.426	-.00125	5.9520	-.03111	-.00311
#3	.00322	-.00277	18.456	.00098	5.9872	-.04457	-.00430
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00262**	**.00921**	F **-.07344**
Stddev	.00140	.00009	.18020
%RSD	53.460	.99471	245.36
#1	-.00401	.00912	.10768
#2	-.00261	.00930	-.07530
#3	-.00122	.00920	-.25270
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6211.9**	**37943.**	**2065.9**
Stddev	10.7	171.	36.3
%RSD	.17281	.45038	1.7553
#1	6203.2	38131.	2024.8
#2	6223.9	37902.	2079.8
#3	6208.6	37796.	2093.2

Sample Name: L1606082303 Acquired: 6/23/2016 20:37:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00036	.09100	.00296	.01304	4.1939	-.00007	203.18
Stddev	.00140	.00531	.00327	.00376	.0297	.00006	1.67
%RSD	388.51	5.8358	110.60	28.801	.70801	88.134	.82056
#1	-.00165	.09683	.00075	.00892	4.2189	-.00004	204.69
#2	.00112	.08975	.00671	.01627	4.1611	-.00014	201.39
#3	-.00055	.08643	.00141	.01393	4.2018	-.00002	203.46
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00062	.00605	.00148	.00325	.50161	1.7415	.06042
Stddev	.00049	.00063	.00113	.00080	.03478	.1908	.00296
%RSD	79.462	10.371	76.602	24.650	6.9332	10.958	4.9058
#1	.00115	.00647	.00038	.00343	.46408	1.6330	.05783
#2	.00054	.00533	.00142	.00237	.53275	1.6297	.06365
#3	.00017	.00634	.00265	.00394	.50800	1.9619	.05979
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	118.15	2.3989	.00023	F 640.12	.00507	.01016	.00289
Stddev	.81	.0242	.00018	10.00	.00123	.00431	.00149
%RSD	.68254	1.0089	76.880	1.5615	24.328	42.407	51.612
#1	118.94	2.4148	.00038	650.69	.00487	.00561	.00152
#2	117.33	2.3711	.00027	630.82	.00639	.01069	.00267
#3	118.17	2.4110	.00003	638.86	.00395	.01418	.00448
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Sample Name: L1606082303 Acquired: 6/23/2016 20:37:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00012**	**-.00944**	**13.225**	**-.00084**	**3.8421**	F **-.03122**	**-.00075**
Stddev	.00335	.00541	.023	.00156	.0312	.00884	.00263
%RSD	2724.2	57.378	.17522	185.97	.81266	28.325	351.62
#1	.00362	-.00548	13.247	.00011	3.8716	-.02289	.00226
#2	-.00306	-.00723	13.225	-.00264	3.8094	-.04050	-.00263
#3	-.00019	-.01560	13.201	.00001	3.8453	-.03028	-.00187
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00215**	**.00570**	**.39111**
Stddev	.00161	.00012	.18787
%RSD	74.559	2.0758	48.034
#1	-.00169	.00568	.30759
#2	-.00394	.00583	.60626
#3	-.00083	.00559	.25948
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6278.4**	**38565.**	**2078.5**
Stddev	22.0	175.	41.0
%RSD	.35101	.45432	1.9726
#1	6265.3	38533.	2033.9
#2	6266.0	38754.	2114.5
#3	6303.8	38408.	2087.2

Sample Name: L1606082304 Acquired: 6/23/2016 20:41:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00041**	**.00982**	**.00065**	**.03078**	**1.3316**	**-.00001**	**46.431**
Stddev	.00324	.01473	.00404	.00488	.0087	.00012	.330
%RSD	781.99	150.01	619.18	15.855	.65651	2204.3	.71076
#1	.00226	.02597	.00499	.03065	1.3415	-.00014	46.804
#2	.00231	.00636	-.00301	.03572	1.3252	.00009	46.176
#3	-.00333	-.00287	-.00002	.02596	1.3280	.00003	46.315
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00066**	**-.00016**	F **-.00347**	**.00119**	**84.345**	**3.2107**	**.04191**
Stddev	.00021	.00040	.00039	.00149	.339	.1063	.00251
%RSD	31.873	248.01	11.380	125.50	.40248	3.3107	5.9779
#1	.00060	-.00031	-.00306	.00290	84.661	3.3331	.04445
#2	.00048	-.00046	-.00385	.00044	84.387	3.1411	.03944
#3	.00089	.00029	-.00349	.00022	83.986	3.1580	.04185
Check ?	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit			36.000				
Low Limit			-.00300				

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**29.094**	**1.4706**	**-.00067**	**141.52**	**.00104**	**.07644**	**.00312**
Stddev	.106	.0113	.00040	1.02	.00036	.00458	.00219
%RSD	.36352	.76758	59.671	.72353	34.222	5.9922	69.972
#1	29.216	1.4831	-.00079	142.69	.00136	.08096	.00145
#2	29.030	1.4675	-.00100	141.08	.00110	.07654	.00233
#3	29.037	1.4612	-.00023	140.78	.00066	.07180	.00560
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606082304 Acquired: 6/23/2016 20:41:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**-.00444**	**16.946**	**-.00016**	**1.0594**	**.01618**	**-.00009**
Stddev	.00138	.00283	.014	.00045	.0066	.00613	.00273
%RSD	1084.7	63.725	.08315	282.51	.62739	37.902	3068.2
#1	-.00159	-.00414	16.934	.00035	1.0669	.01325	.00044
#2	.00005	-.00741	16.943	-.00032	1.0545	.02323	.00234
#3	.00116	-.00177	16.961	-.00051	1.0567	.01207	-.00305
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00431**	**.01323**	F **-.28757**
Stddev	.00070	.00016	.06453
%RSD	16.230	1.2047	22.439
#1	-.00354	.01326	-.33446
#2	-.00489	.01337	-.31427
#3	-.00452	.01306	-.21398
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6548.1**	**40764.**	**2104.8**
Stddev	18.7	118.	30.0
%RSD	.28493	.28837	1.4234
#1	6562.8	40898.	2079.6
#2	6554.4	40678.	2096.8
#3	6527.1	40717.	2137.9

Sample Name: L1606088302 Acquired: 6/23/2016 20:45:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00335	.84495	.01492	.13705	.35776	-.00016	61.342	.00073
Stddev	.00192	.00536	.00123	.00168	.00649	.00001	.914	.00037
%RSD	57.476	.63420	8.2502	1.2235	1.8144	6.0991	1.4894	50.484
#1	.00126	.85090	.01399	.13791	.36483	-.00016	62.379	.00100
#2	.00372	.84051	.01632	.13512	.35638	-.00017	60.987	.00089
#3	.00506	.84344	.01445	.13813	.35207	-.00015	60.658	.00031
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00790	.00132	.00561	5.6286	11.735	.00962	26.311	2.7437
Stddev	.00012	.00055	.00131	.0909	.087	.00112	.394	.0396
%RSD	1.4821	41.328	23.360	1.6145	.74143	11.624	1.4976	1.4443
#1	.00790	.00105	.00615	5.6698	11.835	.00907	26.689	2.7892
#2	.00779	.00096	.00657	5.6916	11.681	.00887	26.343	2.7252
#3	.00802	.00195	.00412	5.5244	11.688	.01090	25.902	2.7167
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01337	55.014	.02594	.07391	.01768	.00363	-.00148	7.7740
Stddev	.00021	.737	.00058	.00987	.00156	.00456	.00201	.0095
%RSD	1.5352	1.3397	2.2195	13.356	8.8413	125.65	135.81	.12214
#1	.01329	55.834	.02652	.06415	.01943	.00884	.00061	7.7638
#2	.01322	54.801	.02594	.08389	.01641	.00164	-.00340	7.7755
#3	.01361	54.406	.02537	.07370	.01720	.00040	-.00164	7.7826
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606088302 Acquired: 6/23/2016 20:45:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00090**	**.36986**	**.01540**	**-.00118**	**.00118**	**.12530**	**1.4316**
Stddev	.00025	.00460	.01415	.00177	.00035	.00030	.3050
%RSD	27.990	1.2450	91.875	150.26	29.961	.24276	21.302
#1	.00061	.37482	.00658	-.00313	.00110	.12559	1.0819
#2	.00104	.36904	.00789	.00033	.00156	.12533	1.6421
#3	.00105	.36572	.03171	-.00073	.00087	.12498	1.5707
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**7921.9**	**49318.**	**2693.6**
Stddev	12.3	123.	25.0
%RSD	.15529	.24979	.92714
#1	7908.5	49427.	2667.6
#2	7924.8	49342.	2717.4
#3	7932.6	49184.	2695.7

Sample Name: CCV Acquired: 6/23/2016 20:49:10 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39590	9.9110	.39526	.48909	.98565	.04996	9.8451
Stddev	.00120	.0339	.00042	.00457	.02803	.00042	.3357
%RSD	.30353	.34216	.10738	.93432	2.8436	.84013	3.4103
#1	.39483	9.8747	.39503	.48786	.95368	.04993	9.4641
#2	.39720	9.9418	.39501	.48526	.99729	.04955	9.9733
#3	.39566	9.9167	.39575	.49415	1.0060	.05039	10.098
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05017	.20113	.49806	.50390	3.9193	49.281	.99930
Stddev	.00013	.00026	.00295	.00102	.1281	1.426	.02793
%RSD	.24979	.12923	.59274	.20262	3.2691	2.8935	2.7947
#1	.05003	.20121	.49990	.50316	3.7715	47.645	.96745
#2	.05025	.20083	.49465	.50347	3.9883	49.937	1.0109
#3	.05024	.20133	.49963	.50506	3.9981	50.261	1.0196
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.8643	.49902	1.0085	49.184	.49977	9.8773	.49959
Stddev	.3185	.01772	.0043	1.350	.00010	.0178	.00372
%RSD	3.2291	3.5503	.42423	2.7456	.01960	.18066	.74560
#1	9.4975	.47954	1.0116	47.634	.49976	9.8611	.50047
#2	10.025	.50333	1.0103	49.812	.49987	9.8744	.49550
#3	10.071	.51418	1.0036	50.106	.49968	9.8964	.50280
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 20:49:10 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2048**	**.40490**	**5.0297**	**1.0004**	**.98725**	**.99313**	**.50297**
Stddev	.0042	.00796	.0039	.0034	.02832	.03209	.00363
%RSD	.34965	1.9650	.07741	.34428	2.8681	3.2307	.72121
#1	1.2087	.41298	5.0259	.99890	.95492	.95803	.50669
#2	1.2003	.40465	5.0296	.99795	.99923	1.0004	.49944
#3	1.2055	.39708	5.0337	1.0043	1.0076	1.0209	.50278
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.99558**	**1.0009**	F **.02289**
Stddev	.00580	.0022	.25371
%RSD	.58246	.21707	1108.2
#1	.99450	.99842	.14548
#2	.99040	1.0017	-.26883
#3	1.0018	1.0025	.19203
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6660.9**	**41801.**	**2169.0**
Stddev	32.3	28.	115.6
%RSD	.48534	.06597	5.3274
#1	6698.0	41774.	2302.5
#2	6639.4	41798.	2102.0
#3	6645.2	41829.	2102.6

Sample Name: CCB Acquired: 6/23/2016 20:52:47 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00186	-.00093	.00316	.00230	.00074	.00006	.02069
Stddev	.00123	.00851	.00342	.00122	.00124	.00008	.05541
%RSD	66.210	911.24	108.19	52.892	166.55	142.19	267.79
#1	.00059	-.01042	.00649	.00167	.00054	-.00003	-.00744
#2	.00195	.00160	.00333	.00153	-.00038	.00006	-.01500
#3	.00305	.00602	-.00034	.00370	.00206	.00014	.08452
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00002	.00039	.00115	.00076	-.01535	.06062	-.00386
Stddev	.00025	.00015	.00083	.00156	.01904	.09913	.00805
%RSD	1273.3	38.748	71.888	205.02	124.01	163.52	208.55
#1	.00015	.00037	.00104	.00077	-.03627	-.05039	.00540
#2	-.00027	.00056	.00039	.00232	.00098	.09196	-.00780
#3	.00017	.00026	.00203	-.00080	-.01077	.14030	-.00918
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01451	-.00216	.00311	.14026	-.00019	.00962	.00081
Stddev	.08285	.00307	.00109	.02230	.00100	.00497	.00347
%RSD	570.85	141.96	34.905	15.899	521.75	51.686	430.66
#1	-.05222	-.00061	.00206	.16585	-.00076	.01507	.00163
#2	.10725	-.00570	.00305	.12497	-.00078	.00533	-.00300
#3	-.01149	-.00018	.00423	.12997	.00096	.00846	.00380
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 20:52:47 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00098**	**-.00415**	**.00357**	**.00020**	**.00045**	**-.00310**	**-.00314**
Stddev	.00119	.00276	.00144	.00034	.00013	.00169	.00399
%RSD	121.00	66.464	40.393	168.69	30.219	54.451	127.02
#1	-.00049	-.00098	.00477	.00034	.00060	-.00134	-.00058
#2	-.00234	-.00542	.00396	-.00019	.00034	-.00325	-.00773
#3	-.00012	-.00603	.00197	.00045	.00040	-.00471	-.00111
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00114**	**.00020**	F **.30264**
Stddev	.00114	.00013	.32767
%RSD	99.237	64.632	108.27
#1	-.00245	.00019	.39170
#2	-.00061	.00033	-.06036
#3	-.00037	.00008	.57657
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6711.0**	**41762.**	**2107.6**
Stddev	39.0	171.	14.5
%RSD	.58138	.41024	.68723
#1	6710.2	41729.	2123.6
#2	6672.4	41947.	2095.3
#3	6750.5	41609.	2103.9

Sample Name: L1606088401 Acquired: 6/23/2016 20:56:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00138**	**.01701**	**.00493**	**.04852**	**.00618**	**-.00007**	**12.628**	**.00247**
Stddev	.00005	.00414	.00116	.00188	.00077	.00012	.147	.00032
%RSD	3.9151	24.333	23.511	3.8697	12.414	161.67	1.1647	12.874
#1	-.00141	.01981	.00570	.04698	.00662	.00006	12.792	.00280
#2	-.00141	.01895	.00549	.04798	.00529	-.00016	12.508	.00217
#3	-.00131	.01225	.00360	.05061	.00662	-.00013	12.584	.00245
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.33204**	**-.00005**	**.12713**	**.00004**	**.29521**	**1.5328**	**3.2466**	**.01448**
Stddev	.00034	.00102	.00048	.01339	.06956	.0143	.1805	.00109
%RSD	.10335	2250.3	.37814	32463.	23.564	.93016	5.5601	7.5480
#1	.33185	.00032	.12750	.00310	.34102	1.5490	3.4388	.01457
#2	.33183	.00074	.12659	-.01461	.32944	1.5271	3.0806	.01335
#3	.33243	-.00120	.12731	.01164	.21517	1.5222	3.2204	.01553
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00093**	**152.98**	**10.187**	**.09320**	**-.00036**	**-.00902**	**-.00316**	**.51224**
Stddev	.00032	1.78	.001	.01280	.00229	.00381	.00857	.00180
%RSD	34.564	1.1612	.01164	13.735	636.44	42.238	271.43	.35176
#1	.00071	155.02	10.189	.07854	-.00010	-.00788	-.00504	.51429
#2	.00079	151.86	10.188	.09884	-.00276	-.00591	.00620	.51148
#3	.00130	152.04	10.186	.10221	.00179	-.01327	-.01063	.51094
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606088401 Acquired: 6/23/2016 20:56:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00055	.06978	-.00044	-.00366	-.00050	.00520	.35447
Stddev	.00021	.00053	.00665	.00240	.00048	.00018	.13127
%RSD	37.225	.76020	1503.3	65.540	95.032	3.3899	37.033
#1	.00077	.07037	-.00450	-.00503	-.00086	.00501	.20475
#2	.00053	.06936	-.00405	-.00089	-.00069	.00524	.40886
#3	.00036	.06960	.00723	-.00506	.00004	.00536	.44980
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6642.8	41062.	2131.9
Stddev	24.2	57.	30.4
%RSD	.36471	.13951	1.4260
#1	6639.2	41009.	2097.0
#2	6620.6	41122.	2152.2
#3	6668.6	41053.	2146.6

Sample Name: L1606089601 Acquired: 6/23/2016 21:00:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00276	.06592	.00541	.02208	.05123	-.00005	4.5311	.00038
Stddev	.00137	.00695	.00371	.00073	.00110	.00012	.0222	.00020
%RSD	49.858	10.542	68.509	3.3035	2.1551	257.77	.49075	51.621
#1	.00357	.06862	.00418	.02271	.05250	.00006	4.5567	.00056
#2	.00353	.05803	.00958	.02226	.05063	-.00017	4.5198	.00017
#3	.00117	.07112	.00247	.02128	.05055	-.00003	4.5168	.00042
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.06787	-.00161	.00146	42.996	.90153	-.00084	5.1151	6.0536
Stddev	.00044	.00070	.00153	.463	.08378	.00102	.1337	.0524
%RSD	.64651	43.319	104.57	1.0779	9.2930	121.94	2.6141	.86472
#1	.06763	-.00220	.00017	43.521	.97406	.00034	5.2374	6.1132
#2	.06761	-.00084	.00314	42.825	.80982	-.00134	4.9723	6.0323
#3	.06838	-.00179	.00107	42.642	.92069	-.00151	5.1355	6.0152
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00036	19.350	.00357	-.00356	.00111	.00096	-.00499	4.5346
Stddev	.00020	.125	.00141	.00838	.00156	.00212	.00300	.0153
%RSD	55.732	.64532	39.385	235.57	139.61	221.63	60.052	.33782
#1	-.00054	19.492	.00220	-.01302	-.00064	.00306	-.00344	4.5188
#2	-.00040	19.300	.00350	.00291	.00231	-.00118	-.00845	4.5355
#3	-.00014	19.259	.00500	-.00056	.00168	.00099	-.00309	4.5494
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089601 Acquired: 6/23/2016 21:00:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00060**	**.03780**	**.00811**	**-.00381**	**-.00115**	**.00560**	**.10083**
Stddev	.00066	.00049	.00739	.00382	.00028	.00007	.23956
%RSD	110.76	1.2963	91.111	100.26	24.644	1.2642	237.60
#1	.00016	.03802	.00291	-.00026	-.00147	.00553	-.17268
#2	-.00087	.03815	.00485	-.00331	-.00104	.00559	.20174
#3	-.00109	.03724	.01658	-.00784	-.00094	.00567	.27342
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6699.7**	**41605.**	**2147.2**
Stddev	6.0	395.	20.8
%RSD	.08901	.94822	.96897
#1	6705.8	41196.	2129.6
#2	6699.3	41634.	2141.8
#3	6693.9	41984.	2170.2

Sample Name: L1606089602 Acquired: 6/23/2016 21:04:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00115	.04083	.00060	.00715	.00811	-.00006	.41120	.00010
Stddev	.00092	.00909	.00372	.00237	.00130	.00006	.07420	.00010
%RSD	79.895	22.273	617.42	33.111	16.069	103.37	18.045	95.174
#1	.00172	.03036	-.00348	.00464	.00961	-.00013	.44439	.00004
#2	.00163	.04672	.00380	.00935	.00726	-.00004	.46300	.00021
#3	.00009	.04541	.00149	.00745	.00746	-.00001	.32619	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00002	.00085	.00131	.07554	1.8076	-.00047	.25019	.00154
Stddev	.00047	.00183	.00071	.02271	.0472	.00199	.07424	.00073
%RSD	2125.6	214.06	54.320	30.062	2.6112	424.18	29.673	47.627
#1	-.00052	.00002	.00060	.10056	1.7705	.00136	.32911	.00152
#2	.00004	.00295	.00130	.05624	1.7915	-.00018	.18174	.00229
#3	.00042	-.00040	.00202	.06981	1.8607	-.00259	.23972	.00082
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00001	8.3070	.00081	-.00279	.00077	-.00363	-.00191	9.3341
Stddev	.00042	.0355	.00152	.00730	.00368	.00212	.00752	.0120
%RSD	4219.9	.42734	186.83	261.51	478.27	58.376	393.78	.12811
#1	-.00026	8.3438	.00114	-.00340	.00064	-.00604	.00678	9.3471
#2	.00047	8.2730	-.00084	.00479	-.00284	-.00208	-.00623	9.3318
#3	-.00024	8.3042	.00214	-.00976	.00451	-.00276	-.00628	9.3235
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089602 Acquired: 6/23/2016 21:04:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00104**	**.00271**	**.00141**	**-.00441**	**.00055**	**.00294**	**.09481**
Stddev	.00076	.00065	.00707	.00096	.00059	.00032	.12117
%RSD	73.387	23.817	499.85	21.749	106.89	10.896	127.81
#1	-.00017	.00268	.00405	-.00333	.00004	.00305	.13413
#2	-.00159	.00337	.00678	-.00473	.00120	.00257	-.04114
#3	-.00135	.00208	-.00659	-.00517	.00043	.00318	.19143
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6686.8**	**41808.**	**2134.5**
Stddev	18.1	213.	20.0
%RSD	.27005	.50969	.93843
#1	6699.2	41631.	2111.6
#2	6666.1	41748.	2148.9
#3	6695.1	42044.	2142.9

Sample Name: L1606089603 Acquired: 6/23/2016 21:08:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00102	.11801	.00474	.00963	.06515	.00009	2.8598	-.00002
Stddev	.00304	.00302	.00162	.00102	.00099	.00007	.0117	.00049
%RSD	298.48	2.5577	34.279	10.645	1.5207	83.968	.40938	2310.2
#1	.00427	.12112	.00606	.01075	.06530	.00010	2.8502	-.00015
#2	.00053	.11781	.00292	.00939	.06607	.00015	2.8729	-.00043
#3	-.00175	.11509	.00524	.00874	.06410	.00001	2.8564	.00052
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00155	-.00024	.00127	2.7358	3.5730	.00040	1.7276	.02916
Stddev	.00022	.00132	.00125	.0173	.1870	.00308	.0541	.00143
%RSD	13.914	551.57	98.352	.63220	5.2335	762.16	3.1323	4.9029
#1	.00170	-.00168	.00263	2.7557	3.7847	.00309	1.6947	.02869
#2	.00164	.00005	.00099	2.7247	3.4302	-.00295	1.7900	.03077
#3	.00130	.00091	.00018	2.7268	3.5042	.00107	1.6980	.02803
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00031	7.6284	.00235	.00047	-.00082	.00245	-.00905	5.4943
Stddev	.00032	.0301	.00079	.01031	.00433	.00306	.00148	.0133
%RSD	102.08	.39457	33.547	2191.2	526.63	125.12	16.371	.24152
#1	-.00063	7.6521	.00324	.00549	-.00579	.00459	-.00977	5.4890
#2	-.00033	7.6386	.00175	-.01139	.00215	-.00106	-.01004	5.4845
#3	.00001	7.5946	.00205	.00731	.00118	.00382	-.00735	5.5094
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089603 Acquired: 6/23/2016 21:08:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00012	.02218	.00164	-.00426	-.00004	.00835	.14767
Stddev	.00040	.00013	.00288	.00161	.00009	.00018	.44819
%RSD	344.86	.60531	175.60	37.721	191.36	2.1138	303.52
#1	.00047	.02205	-.00008	-.00574	-.00012	.00836	-.32337
#2	-.00031	.02232	.00496	-.00255	.00005	.00853	.19752
#3	.00019	.02218	.00004	-.00448	-.00006	.00817	.56885
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6682.9	41728.	2128.8
Stddev	17.5	288.	26.5
%RSD	.26233	.68938	1.2452
#1	6675.1	41447.	2098.3
#2	6670.6	42022.	2146.0
#3	6703.0	41715.	2142.2

Sample Name: L1606089604 Acquired: 6/23/2016 21:12:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00189	.07787	.00274	.02534	.04671	-.00007	4.6127
Stddev	.00185	.01361	.00115	.00174	.00203	.00009	.0084
%RSD	97.576	17.475	41.979	6.8483	4.3441	132.30	.18280
#1	.00394	.07337	.00237	.02334	.04637	.00002	4.6224
#2	.00037	.09315	.00403	.02650	.04487	-.00016	4.6067
#3	.00136	.06707	.00182	.02616	.04888	-.00006	4.6091
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00046	.07375	-.00139	.00045	44.265	.94195	-.00388
Stddev	.00010	.00022	.00095	.00149	.255	.04262	.00409
%RSD	21.656	.29660	68.674	332.55	.57666	4.5243	105.62
#1	.00034	.07398	-.00166	-.00009	44.524	.94561	-.00857
#2	.00051	.07354	-.00218	-.00070	44.257	.89762	-.00204
#3	.00052	.07372	-.00033	.00214	44.014	.98262	-.00102
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	6.3790	6.1215	-.00097	19.253	.00495	-.00232	.00060
Stddev	.0237	.0305	.00040	.056	.00026	.00993	.00262
%RSD	.37160	.49866	40.991	.29188	5.2685	429.04	440.43
#1	6.3989	6.1509	-.00053	19.318	.00524	-.00316	-.00241
#2	6.3528	6.1237	-.00131	19.216	.00473	-.01180	.00179
#3	6.3853	6.0899	-.00108	19.227	.00489	.00801	.00241
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606089604 Acquired: 6/23/2016 21:12:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00465**	**-.00637**	**4.5161**	**-.00039**	**.04033**	**.00878**	**-.00291**
Stddev	.00319	.00325	.0073	.00021	.00012	.01173	.00069
%RSD	68.469	51.078	.16173	52.620	.29003	133.62	23.851
#1	.00106	-.00941	4.5223	-.00056	.04037	.01678	-.00220
#2	.00578	-.00675	4.5081	-.00016	.04019	.01426	-.00359
#3	.00712	-.00294	4.5179	-.00046	.04041	-.00469	-.00294
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00260**	**.00552**	F **-.09796**
Stddev	.00057	.00025	.21094
%RSD	21.873	4.5503	215.32
#1	-.00325	.00580	.14398
#2	-.00238	.00531	-.24325
#3	-.00218	.00544	-.19462
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6689.8**	**41666.**	**2101.2**
Stddev	9.9	402.	24.3
%RSD	.14842	.96487	1.1576
#1	6695.2	41372.	2073.4
#2	6678.4	42124.	2111.3
#3	6695.9	41503.	2118.8

Sample Name: L1606089605 Acquired: 6/23/2016 21:16:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00006	.04830	-.00145	.00590	.06272	.00007	2.6789	.00041
Stddev	.00176	.01001	.00427	.00202	.00179	.00010	.0329	.00039
%RSD	2993.4	20.730	295.15	34.254	2.8603	144.40	1.2294	96.299
#1	-.00190	.05947	.00328	.00605	.06424	-.00005	2.7135	.00011
#2	.00153	.04527	-.00502	.00784	.06318	.00013	2.6479	.00085
#3	.00055	.04014	-.00260	.00381	.06074	.00012	2.6754	.00025
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00157	-.00035	.00184	.89250	3.4230	-.00452	1.6804	.01911
Stddev	.00028	.00023	.00080	.02843	.0162	.00117	.0608	.00353
%RSD	17.515	67.269	43.308	3.1854	.47175	25.991	3.6163	18.451
#1	.00131	-.00060	.00162	.86545	3.4125	-.00336	1.6187	.02198
#2	.00155	-.00014	.00273	.92213	3.4148	-.00448	1.7402	.01517
#3	.00186	-.00031	.00117	.88990	3.4416	-.00571	1.6824	.02016
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00024	7.0650	.00185	-.00101	-.00247	.00326	-.00672	5.2164
Stddev	.00018	.0411	.00050	.00181	.00243	.00063	.00650	.0100
%RSD	75.478	.58180	27.196	179.64	98.576	19.406	96.645	.19223
#1	-.00033	7.1107	.00239	-.00299	-.00232	.00293	.00077	5.2089
#2	-.00037	7.0530	.00139	.00057	-.00497	.00286	-.01006	5.2125
#3	-.00003	7.0312	.00176	-.00061	-.00011	.00399	-.01087	5.2278
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089605 Acquired: 6/23/2016 21:16:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00041	.02070	-.00116	-.00495	-.00050	.00647	.22434
Stddev	.00061	.00025	.00062	.00270	.00039	.00006	.14133
%RSD	147.06	1.2106	53.841	54.602	79.342	1.0037	62.996
#1	-.00061	.02054	-.00094	-.00697	-.00093	.00640	.33918
#2	.00027	.02099	-.00186	-.00188	-.00039	.00649	.26733
#3	-.00089	.02057	-.00067	-.00601	-.00017	.00652	.06651
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6679.9	41653.	2126.5
Stddev	8.5	193.	24.2
%RSD	.12769	.46319	1.1397
#1	6686.1	41709.	2098.6
#2	6670.2	41811.	2142.5
#3	6683.5	41438.	2138.4

Sample Name: L1606089607 Acquired: 6/23/2016 21:20:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00019**	**.00633**	**.00159**	**.00141**	**-.00044**	**-.00001**	**.07933**
Stddev	.00195	.00666	.00483	.00232	.00080	.00015	.08022
%RSD	1049.3	105.18	304.39	163.83	179.45	1810.1	101.12
#1	-.00175	.01203	-.00042	.00047	-.00070	-.00018	.13178
#2	.00017	.00795	.00710	-.00028	-.00108	.00005	-.01301
#3	.00214	-.00099	-.00192	.00405	.00045	.00010	.11922
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00011**	**.00023**	**.00018**	**.00261**	**-.00047**	**.03391**	**-.00323**
Stddev	.00029	.00036	.00083	.00135	.02031	.06552	.00128
%RSD	269.77	155.32	469.03	51.632	4346.1	193.24	39.606
#1	-.00002	.00060	-.00019	.00137	-.00573	.10021	-.00211
#2	-.00010	-.00011	.00113	.00242	.02196	-.03081	-.00295
#3	.00044	.00020	-.00041	.00405	-.01763	.03232	-.00462
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.02721**	**.00109**	**-.00056**	**.23253**	**.00049**	**-.00816**	**-.00015**
Stddev	.04069	.00341	.00026	.02524	.00107	.00624	.00442
%RSD	149.52	313.70	46.647	10.856	219.38	76.495	2897.5
#1	.01942	-.00058	-.00035	.24925	-.00074	-.01380	.00125
#2	-.05550	-.00117	-.00047	.20349	.00124	-.00145	-.00510
#3	-.04556	.00501	-.00085	.24484	.00096	-.00923	.00339
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606089607 Acquired: 6/23/2016 21:20:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00194**	F **-.01015**	**.08416**	**.00022**	**-.00011**	**-.00212**	**-.00416**
Stddev	.00173	.00222	.00063	.00064	.00006	.01168	.00343
%RSD	89.180	21.886	.74271	295.50	49.832	550.35	82.554
#1	-.00101	-.00764	.08452	-.00003	-.00011	-.01340	-.00177
#2	-.00087	-.01186	.08451	-.00027	-.00006	.00993	-.00262
#3	-.00393	-.01095	.08344	.00095	-.00017	-.00289	-.00809
Check ?	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit		90.000					
Low Limit		-.01000					

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00083**	**.00158**	**.28593**
Stddev	.00029	.00007	.07024
%RSD	34.545	4.6737	24.566
#1	-.00055	.00150	.21816
#2	-.00081	.00165	.28123
#3	-.00112	.00159	.35840
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6700.8**	**41736.**	**2133.1**
Stddev	34.6	206.	28.2
%RSD	.51651	.49285	1.3235
#1	6660.9	41521.	2103.9
#2	6723.0	41931.	2160.2
#3	6718.5	41757.	2135.2

Sample Name: L1606082202 Acquired: 6/23/2016 21:24:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00194**	**-.00446**	**.00305**	**.00265**	**.00700**	**.00002**	**.35406**
Stddev	.00134	.00435	.00421	.00519	.00205	.00001	.05184
%RSD	68.787	97.453	137.81	196.12	29.341	37.190	14.641
#1	.00075	-.00690	-.00162	.00667	.00636	.00002	.30481
#2	.00338	.00056	.00423	-.00321	.00930	.00002	.40815
#3	.00169	-.00704	.00655	.00448	.00534	.00003	.34922
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00007**	**.00043**	**-.00077**	**.00285**	**.14848**	**-.04278**	**.00212**
Stddev	.00027	.00044	.00098	.00102	.00792	.02901	.00038
%RSD	394.25	102.19	127.46	35.949	5.3356	67.810	17.947
#1	-.00009	.00000	-.00184	.00239	.15626	-.06821	.00218
#2	.00038	.00041	.00008	.00403	.14874	-.04895	.00171
#3	-.00008	.00089	-.00054	.00214	.14043	-.01118	.00246
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.28223**	**.00480**	**-.00038**	**4.3287**	**-.00014**	**.01612**	**.00046**
Stddev	.10559	.00363	.00060	.1420	.00015	.00579	.00219
%RSD	37.411	75.632	157.86	3.2805	107.54	35.905	474.81
#1	.24611	.00764	-.00099	4.3698	-.00016	.01856	-.00192
#2	.19945	.00071	.00020	4.4456	-.00027	.02028	.00239
#3	.40114	.00605	-.00035	4.1706	.00002	.00951	.00092
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606082202 Acquired: 6/23/2016 21:24:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00001	-.00096	.60016	-.00066	.01034	-.00345	-.00121
Stddev	.00119	.00324	.00463	.00053	.00056	.00260	.00164
%RSD	9663.9	335.88	.77167	79.904	5.4479	75.303	135.44
#1	-.00094	.00170	.59620	-.00020	.00979	-.00115	-.00170
#2	.00134	-.00002	.59904	-.00124	.01092	-.00293	.00062
#3	-.00037	-.00457	.60525	-.00055	.01031	-.00627	-.00255
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.00034	.00082	F -.14423
Stddev	.00011	.00006	.30331
%RSD	33.594	7.1986	210.30
#1	.00036	.00078	-.38581
#2	.00045	.00089	-.24305
#3	.00022	.00079	.19617
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6775.7	42604.	2060.5
Stddev	51.6	45.	84.1
%RSD	.76190	.10471	4.0814
#1	6799.5	42589.	2047.0
#2	6811.2	42654.	1984.0
#3	6716.5	42568.	2150.6

Sample Name: L1606082203 Acquired: 6/23/2016 21:28:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00145	.00184	.00074	-.00019	.03852	.00003	1.5262	.00007
Stddev	.00122	.00469	.00175	.00116	.00077	.00012	.0464	.00017
%RSD	84.264	255.24	237.52	615.42	2.0040	388.98	3.0406	258.31
#1	.00264	-.00346	.00229	-.00083	.03875	.00010	1.5782	-.00003
#2	.00020	.00352	.00109	-.00089	.03915	.00011	1.4889	-.00004
#3	.00151	.00546	-.00116	.00115	.03766	-.00011	1.5116	.00026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00037	.00018	.00108	.25045	2.4536	.00344	.67766	.01002
Stddev	.00037	.00129	.00050	.01436	.0243	.00221	.03143	.00186
%RSD	100.70	701.59	46.418	5.7329	.98953	64.075	4.6384	18.607
#1	.00035	.00006	.00058	.23440	2.4813	.00332	.65654	.01217
#2	.00075	.00153	.00159	.26208	2.4433	.00571	.71379	.00880
#3	.00001	-.00104	.00108	.25488	2.4361	.00130	.66266	.00909
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00035	4.5035	-.00023	.01290	-.00147	.00479	-.00584	.55304
Stddev	.00012	.0716	.00036	.00146	.00189	.00145	.00099	.00147
%RSD	35.728	1.5904	157.06	11.351	128.61	30.358	16.944	.26494
#1	-.00047	4.5590	.00019	.01295	-.00040	.00482	-.00623	.55146
#2	-.00035	4.5290	-.00045	.01141	-.00366	.00623	-.00472	.55435
#3	-.00023	4.4227	-.00042	.01433	-.00036	.00332	-.00658	.55332
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606082203 Acquired: 6/23/2016 21:28:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00002	.03922	-.00354	-.00320	-.00104	.00103	-.02734
Stddev	.00093	.00042	.00540	.00176	.00049	.00021	.22820
%RSD	4945.8	1.0630	152.76	54.899	47.016	20.853	834.62
#1	.00001	.03970	-.00944	-.00516	-.00118	.00102	.23169
#2	.00095	.03900	.00116	-.00265	-.00144	.00082	-.19873
#3	-.00091	.03895	-.00233	-.00178	-.00049	.00125	-.11499
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6687.3	41441.	2081.5
Stddev	25.3	113.	22.3
%RSD	.37870	.27264	1.0699
#1	6658.6	41368.	2073.0
#2	6706.4	41384.	2064.8
#3	6696.9	41571.	2106.8

Sample Name: L1606082302 Acquired: 6/23/2016 21:32:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00042**	**.00297**	**.00273**	**.00225**	**.36313**	**-.00000**	**11.836**
Stddev	.00174	.00449	.00180	.00282	.00235	.00016	.075
%RSD	415.32	151.16	65.994	125.04	.64751	6744.5	.63384
#1	-.00148	.00121	.00168	.00506	.36408	.00008	11.919
#2	.00080	.00807	.00170	.00226	.36486	-.00018	11.773
#3	.00194	-.00038	.00482	-.00057	.36045	.00010	11.815
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00016**	**.00291**	**.00028**	**.00150**	**.22899**	**.13342**	**.00323**
Stddev	.00018	.00030	.00095	.00047	.02084	.11756	.00252
%RSD	113.97	10.305	343.96	31.522	9.1002	88.111	78.086
#1	-.00018	.00302	-.00007	.00192	.21561	.24582	.00054
#2	.00003	.00257	-.00045	.00158	.21836	.01130	.00554
#3	-.00032	.00314	.00135	.00099	.25300	.14315	.00359
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.3954**	**.04077**	**-.00002**	**33.177**	**.00094**	**.00875**	**-.00140**
Stddev	.0055	.00391	.00017	.158	.00081	.00388	.00293
%RSD	.05884	9.5814	818.90	.47682	86.099	44.289	209.14
#1	9.3987	.04528	.00014	33.360	.00025	.00895	-.00476
#2	9.3986	.03829	-.00000	33.085	.00074	.00478	.00055
#3	9.3891	.03876	-.00020	33.087	.00184	.01252	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606082302 Acquired: 6/23/2016 21:32:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00211	F -.01027	.87158	-.00061	.28366	-.00294	-.00367
Stddev	.00335	.00460	.01165	.00033	.00032	.00246	.00356
%RSD	158.22	44.796	1.3368	53.749	.11189	83.646	96.934
#1	.00240	-.01556	.85944	-.00095	.28358	-.00225	-.00028
#2	.00531	-.00804	.87263	-.00029	.28401	-.00568	-.00738
#3	-.00136	-.00721	.88268	-.00060	.28339	-.00091	-.00335
Check ?	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit		90.000					
Low Limit		-.01000					

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	-.00036	.00115	.04430
Stddev	.00028	.00013	.31486
%RSD	77.980	11.657	710.78
#1	-.00058	.00122	.37480
#2	-.00005	.00100	.01024
#3	-.00044	.00124	-.25215
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6666.9	40963.	2080.0
Stddev	54.9	248.	35.1
%RSD	.82321	.60425	1.6856
#1	6605.2	41025.	2039.5
#2	6685.2	40690.	2100.5
#3	6710.3	41173.	2100.0

Sample Name: CCV Acquired: 6/23/2016 21:36:11 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.37886	9.4008	.36977	.46580	.94543	.04748	9.4956
Stddev	.00192	.0240	.00492	.00165	.00279	.00009	.0298
%RSD	.50559	.25544	1.3304	.35366	.29493	.18359	.31411
#1	.37848	9.4279	.37519	.46422	.94745	.04749	9.4832
#2	.38093	9.3822	.36851	.46567	.94658	.04756	9.4739
#3	.37716	9.3923	.36560	.46751	.94225	.04739	9.5296
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04722	.18900	.47404	.47679	3.8147	47.440	.96725
Stddev	.00010	.00048	.00128	.00128	.0064	.112	.00455
%RSD	.21398	.25239	.26979	.26927	.16823	.23504	.47080
#1	.04722	.18886	.47544	.47536	3.8073	47.493	.96637
#2	.04711	.18861	.47292	.47785	3.8188	47.516	.97218
#3	.04731	.18953	.47377	.47715	3.8179	47.312	.96320
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.5275	.47798	.95150	47.323	.47305	9.3519	.47463
Stddev	.0772	.00629	.00356	.158	.00089	.0108	.00368
%RSD	.81075	1.3163	.37366	.33377	.18727	.11544	.77586
#1	9.5741	.48274	.95557	47.363	.47379	9.3402	.47038
#2	9.4383	.47085	.94902	47.457	.47331	9.3538	.47650
#3	9.5701	.48036	.94991	47.149	.47207	9.3616	.47699
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 21:36:11 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1248**	**.37983**	**4.7469**	**.94195**	**.94936**	**.95013**	**.47646**
Stddev	.0071	.00577	.0107	.00244	.00345	.00861	.00052
%RSD	.63160	1.5193	.22491	.25903	.36313	.90637	.10952
#1	1.1176	.38628	4.7372	.93943	.94930	.94189	.47702
#2	1.1251	.37515	4.7451	.94210	.95284	.94943	.47599
#3	1.1318	.37806	4.7583	.94431	.94595	.95907	.47637
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.94937**	**.94698**	F **.53390**
Stddev	.00172	.00064	.57416
%RSD	.18154	.06770	107.54
#1	.95103	.94624	-.12311
#2	.94951	.94740	.93934
#3	.94759	.94730	.78547
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6847.6**	**42330.**	**2179.9**
Stddev	26.3	36.	23.0
%RSD	.38348	.08457	1.0562
#1	6817.2	42321.	2206.4
#2	6862.7	42299.	2165.2
#3	6862.8	42369.	2168.0

Sample Name: CCB Acquired: 6/23/2016 21:39:49 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00034**	**-.00840**	**.00142**	**.00310**	**-.00097**	**-.00004**	**.04289**
Stddev	.00215	.00740	.00154	.00403	.00114	.00009	.02517
%RSD	634.22	88.117	107.88	130.11	117.13	211.10	58.687
#1	.00114	-.00019	.00313	.00134	-.00178	-.00003	.05755
#2	.00065	-.01045	.00097	.00025	-.00148	.00004	.01383
#3	-.00280	-.01457	.00016	.00772	.00033	-.00014	.05730
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00014**	**.00024**	**-.00021**	**.00099**	**.00681**	**.02724**	**.00474**
Stddev	.00017	.00030	.00131	.00197	.01403	.13786	.00231
%RSD	120.69	124.53	625.16	199.14	206.17	506.05	48.675
#1	-.00003	.00049	-.00143	.00036	-.00450	-.09799	.00738
#2	.00014	.00033	-.00037	.00320	.02251	.00476	.00369
#3	.00030	-.00009	.00117	-.00059	.00241	.17496	.00314
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.01627**	**.00349**	**.00316**	**.05656**	**.00129**	**.00259**	**.00220**
Stddev	.06560	.00061	.00054	.03460	.00090	.00126	.00512
%RSD	403.12	17.500	17.006	61.166	69.213	48.482	232.87
#1	.04308	.00418	.00337	.06648	.00145	.00133	.00302
#2	-.00519	.00302	.00356	.01809	.00033	.00384	-.00328
#3	-.08670	.00326	.00255	.08512	.00210	.00261	.00686
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 21:39:49 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00210**	**.00060**	**-.00131**	**-.00025**	**.00045**	**.00101**	**-.00537**
Stddev	.00282	.00202	.00149	.00144	.00024	.00715	.00270
%RSD	134.10	334.37	113.55	585.99	53.933	705.65	50.282
#1	.00467	-.00103	-.00073	-.00138	.00029	-.00721	-.00765
#2	-.00091	.00286	-.00300	-.00074	.00033	.00583	-.00608
#3	.00255	-.00002	-.00020	.00138	.00073	.00442	-.00239
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00036**	**.00019**	F **.41711**
Stddev	.00059	.00027	.26410
%RSD	164.24	139.43	63.317
#1	-.00011	-.00011	.71541
#2	.00007	.00039	.21306
#3	-.00103	.00030	.32286
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6996.3**	**43082.**	**2193.8**
Stddev	20.8	164.	36.3
%RSD	.29721	.37998	1.6551
#1	6972.3	42895.	2225.5
#2	7007.7	43194.	2201.8
#3	7008.9	43159.	2154.2

Sample Name: L1606082303 Acquired: 6/23/2016 21:43:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00032	.00528	-.00158	.00205	.19615	.00000	9.5412	-.00005
Stddev	.00172	.00389	.00374	.00056	.00167	.00004	.0635	.00015
%RSD	531.26	73.680	236.25	27.382	.84980	2856.4	.66503	295.12
#1	.00039	.00192	-.00320	.00177	.19699	.00004	9.6087	.00011
#2	.00201	.00955	-.00424	.00168	.19423	-.00003	9.4828	-.00019
#3	-.00143	.00438	.00269	.00269	.19724	-.00001	9.5322	-.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	-.00003	.00137	.01816	.07360	-.00201	5.7035	.11460
Stddev	.00028	.00217	.00077	.02760	.04149	.00162	.1926	.00292
%RSD	57.802	7719.6	56.083	151.98	56.373	80.472	3.3772	2.5499
#1	.00048	-.00239	.00048	.04879	.11308	-.00147	5.9188	.11188
#2	.00020	.00186	.00178	-.00477	.07737	-.00383	5.5474	.11423
#3	.00076	.00045	.00184	.01046	.03036	-.00073	5.6443	.11769
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00025	30.457	.00028	.01105	.00022	.00112	-.00912	.60825
Stddev	.00054	.348	.00034	.00162	.00227	.00456	.00372	.00164
%RSD	210.59	1.1437	122.90	14.679	1048.7	408.61	40.764	.26908
#1	-.00033	30.637	-.00010	.01036	-.00046	-.00379	-.00828	.60645
#2	.00038	30.056	.00057	.01291	.00274	.00192	-.00590	.60965
#3	.00071	30.679	.00036	.00990	-.00164	.00522	-.01319	.60865
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606082303 Acquired: 6/23/2016 21:43:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00070**	**.18364**	**-.00297**	**-.00161**	**.00010**	**.00105**	**.18385**
Stddev	.00086	.00144	.00537	.00137	.00034	.00025	.28076
%RSD	122.13	.78353	180.76	84.915	342.41	23.814	152.71
#1	-.00168	.18402	-.00642	-.00247	-.00015	.00082	-.00761
#2	-.00010	.18205	.00322	-.00003	-.00005	.00103	.50616
#3	-.00032	.18485	-.00570	-.00232	.00050	.00132	.05301
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6737.6**	**41321.**	**2154.5**
Stddev	10.1	398.	10.5
%RSD	.15056	.96263	.48515
#1	6747.5	41745.	2162.1
#2	6727.2	41262.	2142.5
#3	6738.0	40956.	2158.7

Sample Name: L1606082304 Acquired: 6/23/2016 21:47:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00107	.00763	.00205	.00128	.06210	-.00005	2.2162	-.00002
Stddev	.00070	.00357	.00361	.00281	.00225	.00003	.0545	.00005
%RSD	65.790	46.743	175.75	220.17	3.6308	56.826	2.4590	342.73
#1	.00064	.01174	.00606	-.00197	.06015	-.00004	2.1625	-.00007
#2	.00188	.00577	-.00095	.00288	.06159	-.00009	2.2715	.00004
#3	.00068	.00537	.00106	.00292	.06457	-.00003	2.2145	-.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	-.00139	.00241	4.0573	.28730	.00089	1.4304	.07372
Stddev	.00015	.00065	.00132	.0516	.13779	.00299	.0553	.00403
%RSD	30.727	46.428	54.848	1.2708	47.960	334.26	3.8623	5.4701
#1	.00066	-.00136	.00091	4.0586	.13895	.00365	1.4701	.07443
#2	.00040	-.00206	.00293	4.0051	.41127	.00130	1.4537	.07736
#3	.00040	-.00076	.00340	4.1081	.31169	-.00228	1.3673	.06938
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00002	6.7673	.00003	.01064	-.00183	.00117	-.00224	.82027
Stddev	.00026	.0658	.00066	.00541	.00174	.00605	.00594	.00641
%RSD	1414.1	.97251	2005.0	50.852	95.153	515.39	265.01	.78184
#1	-.00017	6.7130	.00068	.01686	-.00367	.00122	-.00440	.81299
#2	-.00008	6.7484	.00006	.00808	-.00161	-.00490	-.00681	.82274
#3	.00031	6.8405	-.00064	.00698	-.00021	.00720	.00448	.82508
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606082304 Acquired: 6/23/2016 21:47:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00062**	**.05041**	**.00372**	**-.00162**	**-.00102**	**.00139**	**.02709**
Stddev	.00054	.00053	.00900	.00223	.00128	.00033	.20037
%RSD	86.087	1.0604	241.81	137.59	125.65	23.588	739.53
#1	-.00016	.04984	-.00362	-.00387	-.00076	.00168	-.08640
#2	-.00121	.05049	.01377	-.00158	-.00240	.00104	.25845
#3	-.00050	.05090	.00102	.00059	.00011	.00145	-.09076
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6801.0**	**42034.**	**2119.8**
Stddev	49.8	142.	31.5
%RSD	.73210	.33806	1.4863
#1	6763.5	42178.	2123.4
#2	6782.1	42032.	2149.4
#3	6857.5	41894.	2086.7

Sample Name: L1606088302 Acquired: 6/23/2016 21:51:41 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00081	.03900	.00448	.00430	.01505	-.00001	2.8996	-.00001
Stddev	.00124	.00075	.00243	.00233	.00060	.00004	.0350	.00023
%RSD	153.65	1.9305	54.189	54.088	3.9770	332.40	1.2074	2343.8
#1	-.00027	.03833	.00184	.00580	.01560	.00003	2.9325	.00014
#2	.00216	.03982	.00662	.00548	.01441	-.00003	2.8628	.00010
#3	.00053	.03886	.00496	.00162	.01514	-.00004	2.9035	-.00027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00051	-.00093	.00225	.26018	.53798	-.00133	1.2540	.13015
Stddev	.00005	.00064	.00096	.01052	.04447	.00237	.1006	.00182
%RSD	10.629	68.468	42.462	4.0434	8.2670	177.47	8.0229	1.3964
#1	.00052	-.00155	.00287	.26853	.50853	-.00040	1.3655	.12853
#2	.00056	-.00028	.00273	.26365	.51627	-.00403	1.2265	.12979
#3	.00045	-.00096	.00115	.24836	.58914	.00042	1.1700	.13212
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00023	2.6639	.00090	.01608	.00052	.00237	-.00249	.36519
Stddev	.00017	.0233	.00047	.00836	.00171	.00049	.00862	.00186
%RSD	72.830	.87410	52.415	51.991	330.96	20.707	346.32	.51020
#1	.00029	2.6639	.00099	.02573	.00209	.00189	.00736	.36326
#2	.00036	2.6406	.00039	.01116	.00078	.00287	-.00614	.36532
#3	.00004	2.6871	.00132	.01135	-.00131	.00234	-.00868	.36698
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606088302 Acquired: 6/23/2016 21:51:41 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00006**	**.01730**	**.00326**	**-.00074**	**-.00020**	**.00699**	**.35536**
Stddev	.00047	.00025	.00678	.00023	.00043	.00024	.51910
%RSD	826.36	1.4568	208.35	30.989	219.10	3.4850	146.08
#1	.00036	.01713	.00871	-.00090	-.00020	.00693	-.23473
#2	-.00056	.01718	-.00434	-.00084	-.00062	.00678	.74156
#3	.00004	.01759	.00541	-.00048	.00024	.00726	.55925
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**7019.0**	**43469.**	**2272.3**
Stddev	25.4	199.	24.4
%RSD	.36132	.45762	1.0740
#1	7020.2	43542.	2267.1
#2	6993.1	43621.	2298.9
#3	7043.7	43244.	2250.9

Sample Name: L1606089601 Acquired: 6/23/2016 21:55:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00420**	**.00523**	**.00334**	**.00237**	**.00219**	**.00008**	**.19980**
Stddev	.00278	.00696	.00261	.00237	.00063	.00004	.02591
%RSD	66.297	133.09	78.109	99.957	28.753	51.900	12.967
#1	.00741	.00603	.00348	.00033	.00147	.00009	.18026
#2	.00260	.01174	.00066	.00182	.00265	.00012	.22918
#3	.00258	-.00210	.00587	.00496	.00244	.00003	.18994
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00015**	**.00302**	**.00023**	**.00052**	**2.0027**	**-.01208**	**-.00040**
Stddev	.00012	.00024	.00052	.00149	.0311	.05532	.00593
%RSD	79.007	7.8027	222.58	284.84	1.5506	458.05	1480.9
#1	-.00011	.00287	.00040	.00177	1.9805	.05152	-.00717
#2	-.00005	.00329	-.00035	-.00112	1.9893	-.03870	.00382
#3	-.00028	.00290	.00064	.00091	2.0382	-.04905	.00215
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.24211**	**.28328**	**-.00023**	**.95184**	**.00021**	**.00495**	**-.00083**
Stddev	.13780	.01030	.00015	.04331	.00103	.00358	.00274
%RSD	56.917	3.6352	65.757	4.5498	490.79	72.184	329.16
#1	.26334	.27250	-.00010	.90187	.00062	.00759	.00169
#2	.09492	.28431	-.00019	.97855	.00097	.00088	-.00043
#3	.36806	.29302	-.00040	.97509	-.00096	.00638	-.00376
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606089601 Acquired: 6/23/2016 21:55:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00174**	**-.00552**	**.20894**	**-.00089**	**.00183**	**.00041**	**-.00355**
Stddev	.00455	.00357	.00152	.00075	.00028	.00397	.00196
%RSD	261.01	64.670	.72600	84.586	15.398	961.73	55.133
#1	-.00696	-.00337	.20831	-.00159	.00215	.00126	-.00263
#2	.00141	-.00964	.20784	-.00009	.00174	.00389	-.00222
#3	.00032	-.00355	.21067	-.00099	.00161	-.00391	-.00580
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00023**	**.00107**	F **-.16350**
Stddev	.00118	.00009	.38488
%RSD	512.00	8.7488	235.40
#1	-.00072	.00105	.00228
#2	-.00109	.00099	.11071
#3	.00111	.00117	-.60348
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6828.9**	**41943.**	**2135.6**
Stddev	63.3	286.	20.3
%RSD	.92751	.68281	.94968
#1	6763.6	42271.	2138.9
#2	6833.1	41741.	2154.0
#3	6890.1	41817.	2113.9

Sample Name: L1606089602 Acquired: 6/23/2016 21:59:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00116	.00173	.00317	.00184	-.00011	.00003	.00624	-.00003
Stddev	.00285	.00778	.00249	.00329	.00125	.00006	.05063	.00023
%RSD	245.09	449.18	78.692	179.21	1183.1	208.66	811.35	713.72
#1	-.00055	.01019	.00053	.00389	-.00026	-.00003	-.04528	.00022
#2	.00445	-.00512	.00349	.00357	-.00127	.00009	.05593	-.00023
#3	-.00041	.00012	.00549	-.00196	.00122	.00003	.00807	-.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00024	.00046	.00165	-.00531	.07219	-.00031	.05667	.00209
Stddev	.00028	.00102	.00160	.03109	.09458	.00388	.05878	.00182
%RSD	116.13	222.95	97.271	585.63	131.01	1232.5	103.73	87.087
#1	.00001	.00001	.00226	.02472	.15298	-.00298	.09347	.00378
#2	.00016	-.00027	-.00017	-.00328	.09542	.00413	.08765	.00016
#3	.00055	.00163	.00285	-.03736	-.03184	-.00209	-.01113	.00234
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00056	.39833	-.00077	.00469	.00112	.00031	-.00922	.46168
Stddev	.00006	.02012	.00058	.00167	.00277	.00100	.00659	.00725
%RSD	11.375	5.0505	75.062	35.563	245.83	319.68	71.432	1.5707
#1	-.00061	.37822	-.00128	.00414	.00250	-.00075	-.01634	.45603
#2	-.00049	.39832	-.00088	.00656	-.00206	.00124	-.00799	.45917
#3	-.00057	.41846	-.00014	.00336	.00293	.00045	-.00334	.46986
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089602 Acquired: 6/23/2016 21:59:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00058**	**.00007**	**-.00373**	**-.00407**	**-.00032**	**.00100**	**-.02035**
Stddev	.00028	.00019	.00957	.00088	.00078	.00036	.45457
%RSD	48.712	259.28	256.71	21.746	243.98	36.577	2234.1
#1	-.00053	.00018	.00727	-.00462	.00006	.00142	-.06540
#2	-.00033	.00018	-.00832	-.00454	.00020	.00080	.45507
#3	-.00089	-.00014	-.01012	-.00305	-.00121	.00077	-.45071
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6764.1**	**42193.**	**2152.3**
Stddev	51.5	150.	58.2
%RSD	.76176	.35597	2.7055
#1	6724.7	42019.	2097.0
#2	6745.2	42282.	2213.0
#3	6822.4	42277.	2146.9

Sample Name: L1606089603 Acquired: 6/23/2016 22:03:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00264**	**.00052**	**.00129**	**.00105**	**.00283**	**.00007**	**.13087**	**-.00015**
Stddev	.00174	.01588	.00262	.00317	.00277	.00004	.02295	.00021
%RSD	65.852	3062.5	203.49	301.32	97.964	60.620	17.533	135.33
#1	.00093	.01257	-.00138	.00463	.00302	.00005	.14125	.00004
#2	.00440	.00646	.00387	-.00005	.00549	.00004	.10457	-.00037
#3	.00259	-.01748	.00137	-.00142	-.00004	.00011	.14679	-.00013
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00048**	**.00024**	**.00128**	**.12772**	**.20105**	**-.00470**	**.12553**	**.00043**
Stddev	.00031	.00150	.00216	.02497	.11408	.00639	.06203	.00287
%RSD	63.789	616.72	168.54	19.548	56.742	136.10	49.418	664.80
#1	.00045	.00194	-.00101	.11916	.09190	.00173	.17340	-.00027
#2	.00019	-.00030	.00156	.10815	.19176	-.00476	.14774	-.00202
#3	.00080	-.00091	.00330	.15584	.31949	-.01106	.05545	.00359
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00003**	**.39363**	**-.00055**	**.01390**	**-.00151**	**.00183**	**-.00632**	**.24727**
Stddev	.00045	.04518	.00102	.00707	.00255	.00239	.00219	.00127
%RSD	1327.8	11.477	186.18	50.889	169.36	131.03	34.571	.51246
#1	-.00019	.43317	-.00107	.01491	-.00330	.00328	-.00397	.24716
#2	.00047	.34439	.00063	.02042	.00142	-.00094	-.00828	.24858
#3	-.00038	.40334	-.00121	.00638	-.00264	.00314	-.00672	.24606
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089603 Acquired: 6/23/2016 22:03:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00047**	**.00125**	**-.01273**	**-.00400**	**-.00003**	**.00095**	**.10691**
Stddev	.00077	.00042	.00925	.00266	.00030	.00011	.21605
%RSD	161.90	33.664	72.615	66.501	1119.1	11.082	202.08
#1	-.00019	.00148	-.00295	-.00574	-.00008	.00101	-.14101
#2	-.00134	.00076	-.01391	-.00094	.00030	.00102	.20680
#3	.00011	.00150	-.02133	-.00533	-.00030	.00083	.25493
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6778.0**	**41534.**	**2155.4**
Stddev	33.3	392.	34.0
%RSD	.49153	.94350	1.5766
#1	6804.2	41394.	2116.9
#2	6740.5	41232.	2181.3
#3	6789.2	41977.	2167.8

Sample Name: L1606089604 Acquired: 6/23/2016 22:07:36 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00159	.01026	.00044	.00222	.00123	.00002	.21290	.00018
Stddev	.00043	.01032	.00191	.00166	.00091	.00006	.03422	.00024
%RSD	27.269	100.60	437.55	74.707	74.454	271.36	16.075	135.58
#1	.00148	.01567	.00246	.00277	.00044	.00008	.21883	-.00010
#2	.00207	.01676	-.00133	.00035	.00223	.00002	.24377	.00035
#3	.00122	-.00164	.00018	.00353	.00102	-.00004	.17610	.00029
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00372	-.00098	.00204	1.9871	.04803	-.00029	.27775	.28079
Stddev	.00019	.00146	.00258	.0500	.09028	.00333	.05841	.00512
%RSD	4.9815	148.68	126.42	2.5186	187.98	1134.6	21.028	1.8227
#1	.00391	-.00254	-.00094	1.9637	-.03733	-.00218	.34159	.28507
#2	.00373	.00037	.00345	1.9529	.03888	.00355	.22701	.27512
#3	.00354	-.00079	.00362	2.0445	.14254	-.00224	.26465	.28217
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00026	.88820	.00047	.00405	-.00241	-.00066	-.00911	.20760
Stddev	.00009	.05397	.00049	.00324	.00083	.00328	.00427	.00263
%RSD	33.170	6.0764	104.78	80.127	34.563	498.25	46.914	1.2688
#1	-.00017	.85983	.00038	.00127	-.00145	.00099	-.00982	.20605
#2	-.00027	.85432	.00100	.00326	-.00287	.00147	-.00453	.20610
#3	-.00035	.95043	.00003	.00761	-.00292	-.00443	-.01299	.21064
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089604 Acquired: 6/23/2016 22:07:36 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**.00180**	**.00146**	**-.00008**	**.00002**	**.00083**	**.01606**
Stddev	.00033	.00028	.00448	.00273	.00107	.00008	.15859
%RSD	645.35	15.485	305.81	3306.5	5321.6	9.5646	987.51
#1	-.00043	.00167	.00571	-.00323	-.00007	.00090	-.04948
#2	.00011	.00212	.00190	.00129	-.00100	.00085	-.09926
#3	.00017	.00161	-.00322	.00169	.00113	.00074	.19692
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6852.5**	**41879.**	**2133.3**
Stddev	55.4	422.	31.4
%RSD	.80848	1.0073	1.4729
#1	6793.1	41503.	2112.7
#2	6861.8	42335.	2169.5
#3	6902.7	41799.	2117.7

Sample Name: L1606089605 Acquired: 6/23/2016 22:11:34 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00325**	**-.00813**	**.00321**	**.00103**	**.00232**	**.00007**	**.13232**
Stddev	.00086	.00834	.00166	.00270	.00015	.00005	.04505
%RSD	26.517	102.62	51.687	262.17	6.5199	65.910	34.048
#1	.00357	-.00015	.00448	.00315	.00234	.00011	.13419
#2	.00227	-.00744	.00133	-.00201	.00246	.00002	.08636
#3	.00390	-.01679	.00380	.00194	.00216	.00008	.17641
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00022**	**.00019**	**-.00021**	**-.00051**	**.02399**	**.13993**	**-.00272**
Stddev	.00013	.00048	.00060	.00074	.02883	.04909	.00167
%RSD	59.092	248.65	278.22	144.43	120.17	35.082	61.414
#1	-.00034	.00000	-.00032	-.00045	-.00884	.19045	-.00464
#2	-.00023	.00074	.00043	-.00128	.04519	.13693	-.00190
#3	-.00008	-.00016	-.00075	.00019	.03563	.09241	-.00161
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.13116**	**-.00087**	**-.00017**	**.35584**	**.00083**	**.00721**	**.00009**
Stddev	.12028	.00028	.00012	.03175	.00083	.00389	.00231
%RSD	91.707	32.079	72.680	8.9233	100.69	53.875	2586.3
#1	.15231	-.00055	-.00032	.37833	.00083	.00320	-.00213
#2	.23946	-.00098	-.00010	.31952	-.00001	.01096	.00248
#3	.00170	-.00108	-.00010	.36967	.00166	.00747	-.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606089605 Acquired: 6/23/2016 22:11:34 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00575**	**.00189**	**.24045**	**-.00028**	**.00101**	**-.00455**	**-.00361**
Stddev	.00192	.00588	.00568	.00092	.00006	.01057	.00467
%RSD	33.402	311.30	2.3620	328.38	6.1198	232.48	129.52
#1	.00788	.00697	.23736	-.00119	.00106	-.01626	-.00722
#2	.00521	.00325	.23698	.00066	.00103	-.00167	-.00528
#3	.00416	-.00455	.24700	-.00032	.00094	.00429	.00167
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00106**	**.00104**	F **-.05759**
Stddev	.00074	.00009	.21862
%RSD	69.542	8.5436	379.61
#1	-.00181	.00101	.17219
#2	-.00033	.00097	-.26300
#3	-.00105	.00114	-.08197
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6837.2**	**42004.**	**2152.4**
Stddev	16.0	173.	9.0
%RSD	.23440	.41281	.41704
#1	6853.2	42014.	2142.1
#2	6821.1	41826.	2156.7
#3	6837.2	42172.	2158.4

Sample Name: L1606089607 Acquired: 6/23/2016 22:15:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00124**	**-.00686**	**.00409**	**.00130**	**-.00066**	**-.00005**	**.03206**	**-.00002**
Stddev	.00132	.01058	.00305	.00213	.00072	.00007	.01681	.00018
%RSD	105.94	154.08	74.558	163.59	109.05	125.85	52.443	856.76
#1	-.00264	-.01401	.00622	.00259	-.00111	-.00005	.04548	-.00001
#2	-.00105	-.01187	.00546	.00246	-.00105	.00001	.01320	.00015
#3	-.00003	.00529	.00060	-.00115	.00017	-.00012	.03751	-.00021
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00031**	**.00074**	**.00061**	**-.00910**	**-.00529**	**-.00407**	**.02444**	**.00176**
Stddev	.00024	.00180	.00036	.03159	.06755	.00591	.05716	.00378
%RSD	78.629	244.31	60.097	347.19	1277.4	145.34	233.90	215.59
#1	.00037	.00281	.00094	.01641	.06833	-.00447	-.04095	.00014
#2	.00051	-.00035	.00022	-.04443	-.01978	-.00977	.04932	-.00095
#3	.00004	-.00025	.00066	.00073	-.06441	.00203	.06494	.00608
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00073**	**.00595**	**.00025**	**.00693**	**-.00226**	**-.00313**	**-.00892**	**-.00746**
Stddev	.00042	.03243	.00102	.00571	.00047	.00121	.00509	.00076
%RSD	57.397	545.12	410.30	82.361	20.886	38.630	57.024	10.142
#1	-.00072	.04302	.00024	.00636	-.00177	-.00362	-.00954	-.00789
#2	-.00116	-.01720	-.00077	.01291	-.00271	-.00175	-.01368	-.00792
#3	-.00032	-.00797	.00128	.00153	-.00229	-.00401	-.00356	-.00659
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089607 Acquired: 6/23/2016 22:15:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00050**	**.00006**	**-.00267**	**-.00165**	**-.00070**	**.00059**	**.18991**
Stddev	.00084	.00062	.00845	.00365	.00135	.00010	.03400
%RSD	167.93	1027.2	316.03	221.52	192.73	16.308	17.901
#1	-.00037	-.00011	.00072	-.00580	.00084	.00058	.15066
#2	-.00140	.00075	-.01229	-.00022	-.00169	.00050	.20942
#3	.00027	-.00046	.00355	.00107	-.00126	.00069	.20967
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6854.9**	**42369.**	**2139.0**
Stddev	37.2	282.	6.9
%RSD	.54331	.66600	.32457
#1	6813.1	42056.	2146.4
#2	6867.1	42446.	2132.6
#3	6884.6	42604.	2138.0

Sample Name: CCV Acquired: 6/23/2016 22:19:31 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.37294	9.2890	.37002	.46325	.93343	.04671	9.3219
Stddev	.00058	.0248	.00029	.00298	.00858	.00002	.0885
%RSD	.15646	.26680	.07810	.64298	.91954	.04347	.94918
#1	.37306	9.3173	.36998	.46028	.92385	.04673	9.2510
#2	.37230	9.2715	.36976	.46623	.94043	.04670	9.2937
#3	.37345	9.2781	.37033	.46323	.93599	.04670	9.4210
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04629	.18580	.46699	.46774	3.7375	46.911	.95278
Stddev	.00023	.00060	.00144	.00134	.0018	.467	.00684
%RSD	.48818	.32069	.30929	.28620	.04690	.99443	.71816
#1	.04654	.18545	.46662	.46619	3.7368	46.377	.94493
#2	.04624	.18649	.46859	.46843	3.7395	47.239	.95591
#3	.04609	.18546	.46577	.46859	3.7362	47.116	.95750
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.2808	.46805	.93442	46.553	.46467	9.1850	.46445
Stddev	.1781	.00201	.00336	.432	.00229	.0141	.00099
%RSD	1.9192	.42908	.36011	.92728	.49257	.15361	.21367
#1	9.0878	.46660	.93610	46.055	.46543	9.1763	.46458
#2	9.4388	.47034	.93661	46.821	.46649	9.2013	.46537
#3	9.3158	.46721	.93054	46.783	.46210	9.1774	.46340
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 22:19:31 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1058**	**.37399**	**4.6590**	**.92793**	**.93602**	**.93884**	**.47136**
Stddev	.0068	.00623	.0143	.00333	.00809	.01829	.00375
%RSD	.61331	1.6668	.30694	.35899	.86439	1.9481	.79529
#1	1.1040	.37143	4.6586	.92611	.92682	.91846	.46721
#2	1.1133	.38110	4.6734	.93177	.94202	.94427	.47237
#3	1.1001	.36945	4.6448	.92590	.93921	.95381	.47450
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.93328**	**.93064**	F **.30272**
Stddev	.00145	.00126	.34844
%RSD	.15486	.13568	115.11
#1	.93249	.92922	.70352
#2	.93240	.93165	.07184
#3	.93495	.93106	.13279
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6994.5**	**42945.**	**2201.9**
Stddev	26.0	472.	17.5
%RSD	.37198	1.0980	.79423
#1	7017.9	42835.	2192.4
#2	6966.5	42539.	2222.1
#3	6999.0	43462.	2191.2

Sample Name: CCB Acquired: 6/23/2016 22:23:10 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00075**	**-.01234**	**.00307**	**.00387**	**-.00152**	**.00003**	**.02196**
Stddev	.00074	.00351	.00195	.00329	.00037	.00010	.03751
%RSD	98.761	28.475	63.502	85.024	24.368	329.31	170.81
#1	.00146	-.01497	.00353	.00757	-.00195	-.00000	.04383
#2	.00082	-.00835	.00475	.00124	-.00126	.00014	.04339
#3	-.00002	-.01370	.00093	.00281	-.00136	-.00005	-.02135
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00007**	**.00029**	**-.00074**	**.00053**	**-.01044**	**.06034**	**-.00158**
Stddev	.00023	.00031	.00153	.00009	.01838	.08449	.00205
%RSD	314.24	106.88	205.57	16.843	176.11	140.03	130.11
#1	-.00010	-.00004	-.00241	.00043	-.01522	.02142	-.00392
#2	-.00029	.00057	-.00042	.00060	.00986	.15728	-.00009
#3	.00017	.00033	.00060	.00054	-.02595	.00232	-.00072
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.02604**	**-.00075**	**.00305**	**.06420**	**.00000**	**.00651**	**-.00362**
Stddev	.05654	.00328	.00069	.00709	.00080	.00504	.00339
%RSD	217.15	436.24	22.462	11.042	53218.	77.432	93.854
#1	.03584	-.00047	.00226	.06569	.00068	.00969	-.00632
#2	-.07502	.00238	.00342	.07043	-.00088	.00913	-.00473
#3	-.03894	-.00416	.00347	.05649	.00021	.00070	.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 22:23:10 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00325	-.00859	-.00081	-.00049	.00016	.00095	-.00269
Stddev	.00288	.00100	.00106	.00098	.00044	.00395	.00131
%RSD	88.832	11.667	130.17	199.49	265.70	416.04	48.575
#1	.00425	-.00769	-.00116	-.00120	.00053	-.00220	-.00394
#2	-.00001	-.00968	.00037	.00062	.00028	-.00034	-.00133
#3	.00549	-.00841	-.00166	-.00089	-.00032	.00538	-.00281
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	-.00046	.00011	F .20521
Stddev	.00084	.00022	.40407
%RSD	184.00	201.71	196.91
#1	.00051	-.00006	.57360
#2	-.00094	.00036	-.22696
#3	-.00095	.00003	.26899
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6996.7	43446.	2222.8
Stddev	42.9	110.	21.9
%RSD	.61271	.25379	.98413
#1	6948.6	43485.	2197.6
#2	7030.8	43322.	2237.5
#3	7010.7	43532.	2233.2

Sample Name: LLCCV Acquired: 6/23/2016 22:27:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00907	.13685	.00722	.07327	.00636	.00136	.38737	.00083
Stddev	.00127	.00548	.00269	.00354	.00121	.00002	.02958	.00013
%RSD	14.057	4.0063	37.324	4.8356	18.957	1.6251	7.6350	15.322
#1	.00768	.13064	.00454	.06952	.00698	.00135	.41482	.00073
#2	.01019	.13890	.00718	.07656	.00497	.00135	.39124	.00078
#3	.00934	.14102	.00993	.07374	.00713	.00139	.35605	.00097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00356	.00351	.00643	.06036	.74512	.07304	.35518	.00670
Stddev	.00034	.00083	.00051	.03616	.18336	.00520	.11515	.00028
%RSD	9.6921	23.706	7.9731	59.906	24.608	7.1243	32.420	4.2320
#1	.00383	.00446	.00589	.02265	.76354	.06708	.45336	.00638
#2	.00367	.00316	.00691	.06369	.91857	.07668	.38374	.00683
#3	.00317	.00291	.00648	.09475	.55324	.07536	.22844	.00690
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00743	.43578	.01531	.74167	.00827	.07927	.00569	.77935
Stddev	.00037	.04956	.00080	.00526	.00366	.00493	.00102	.00095
%RSD	5.0056	11.373	5.2515	.70948	44.188	6.2194	17.916	.12251
#1	.00704	.38130	.01503	.74257	.01154	.08395	.00451	.78008
#2	.00779	.47820	.01469	.73602	.00432	.07413	.00628	.77827
#3	.00745	.44783	.01622	.74642	.00896	.07973	.00628	.77970
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: LLCCV Acquired: 6/23/2016 22:27:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.37348	.03865	.02465	.14094	.00676	.01578	19.333
Stddev	.00058	.00060	.00437	.00098	.00113	.00019	.577
%RSD	.15634	1.5628	17.713	.69785	16.700	1.2102	2.9862
#1	.37294	.03795	.02415	.14073	.00773	.01557	19.532
#2	.37410	.03901	.02924	.14008	.00552	.01583	19.783
#3	.37339	.03899	.02055	.14202	.00705	.01594	18.682
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6996.1	43552.	2208.7
Stddev	29.5	407.	29.7
%RSD	.42198	.93392	1.3436
#1	6976.9	43985.	2211.3
#2	7030.1	43494.	2236.9
#3	6981.2	43178.	2177.8

Sample Name: LLCCV Acquired: 6/23/2016 22:31:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02091	.36245	.02162	.18354	.01867	.00356	.98086
Stddev	.00237	.00484	.00558	.00414	.00095	.00018	.04469
%RSD	11.351	1.3361	25.792	2.2543	5.0969	5.1755	4.5561
#1	.01879	.36503	.01522	.18805	.01837	.00338	.94866
#2	.02045	.35687	.02541	.17992	.01791	.00356	.96204
#3	.02347	.36547	.02424	.18267	.01974	.00374	1.0319
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00183	.00959	.00895	.01155	.18702	1.8760	.19064
Stddev	.00040	.00066	.00025	.00083	.02563	.0398	.00709
%RSD	21.803	6.8545	2.7509	7.2203	13.702	2.1189	3.7192
#1	.00229	.01035	.00908	.01200	.20673	1.9212	.19761
#2	.00159	.00922	.00866	.01059	.15805	1.8607	.18344
#3	.00161	.00920	.00910	.01205	.19628	1.8463	.19086
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.90135	.02189	.01808	1.0102	.03771	1.7925	.01852
Stddev	.03529	.00047	.00019	.0359	.00135	.0154	.00112
%RSD	3.9156	2.1507	1.0770	3.5580	3.5742	.86119	6.0288
#1	.87617	.02240	.01801	.97340	.03783	1.7983	.01932
#2	.88620	.02181	.01830	1.0119	.03900	1.8043	.01725
#3	.94169	.02146	.01794	1.0452	.03631	1.7750	.01900
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: LLCCV Acquired: 6/23/2016 22:31:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19437**	**.03462**	**1.8270**	**.92649**	**.09469**	**.05843**	**.34726**
Stddev	.00460	.00283	.0121	.00858	.00060	.00582	.00417
%RSD	2.3690	8.1675	.66332	.92568	.63058	9.9521	1.1997
#1	.19959	.03615	1.8304	.92739	.09439	.06509	.34780
#2	.19266	.03636	1.8370	.93459	.09431	.05586	.35113
#3	.19087	.03136	1.8135	.91750	.09538	.05435	.34285
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.01800**	**.03876**	F **49.426**
Stddev	.00023	.00022	.593
%RSD	1.2842	.57319	1.2005
#1	.01823	.03881	50.072
#2	.01777	.03895	48.906
#3	.01800	.03851	49.300
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**7011.3**	**43477.**	**2210.1**
Stddev	20.4	40.	14.4
%RSD	.29146	.09218	.65240
#1	7004.2	43507.	2200.5
#2	6995.4	43491.	2226.7
#3	7034.3	43431.	2203.1

Sample Name: PBW 11 Acquired: 6/23/2016 22:35:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00032**	**.00826**	**.00073**	**.00097**	**-.00161**	**.00009**	**.02504**
Stddev	.00435	.00626	.00116	.00380	.00132	.00015	.01704
%RSD	1361.2	75.725	158.33	393.95	82.078	159.17	68.056
#1	-.00441	.01091	-.00035	.00280	-.00305	.00000	.03508
#2	.00122	.01276	.00059	.00350	-.00132	.00026	.03466
#3	.00414	.00112	.00196	-.00341	-.00045	.00001	.00536
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00010**	**-.00009**	**.00011**	**.00179**	**-.00443**	**.01747**	**-.00078**
Stddev	.00037	.00030	.00107	.00110	.02744	.11206	.00643
%RSD	372.56	345.12	981.79	61.211	619.83	641.53	826.74
#1	.00053	-.00038	.00079	.00288	.01410	-.01578	-.00199
#2	-.00009	.00021	-.00112	.00069	-.03595	.14238	.00617
#3	-.00014	-.00009	.00066	.00181	.00857	-.07420	-.00651
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.12116**	**.00145**	**-.00010**	**-.00638**	**-.00006**	**-.00844**	**-.00156**
Stddev	.07020	.00190	.00047	.05241	.00004	.00883	.00308
%RSD	57.940	131.08	475.47	821.85	65.273	104.67	197.22
#1	.10625	-.00074	-.00051	-.02102	-.00009	.00126	.00107
#2	.19761	.00240	-.00019	.05180	-.00002	-.01056	-.00495
#3	.05961	.00268	.00041	-.04991	-.00007	-.01601	-.00081
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: PBW 11 Acquired: 6/23/2016 22:35:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-02

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00235**	F **-.01321**	**-.01211**	**.00005**	**.00012**	**-.01090**	**-.00481**
Stddev	.00434	.00615	.00423	.00130	.00022	.00717	.00224
%RSD	184.96	46.568	34.938	2403.9	182.36	65.767	46.622
#1	-.00650	-.00846	-.00872	-.00069	-.00007	-.00262	-.00253
#2	.00215	-.02016	-.01076	-.00070	.00007	-.01519	-.00488
#3	-.00268	-.01101	-.01685	.00155	.00036	-.01489	-.00701
Check ?	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit		90.000					
Low Limit		-.01000					

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00056**	**.00115**	F **-.06214**
Stddev	.00033	.00011	.40590
%RSD	58.758	9.4676	653.23
#1	-.00093	.00115	.04499
#2	-.00040	.00127	.27945
#3	-.00033	.00105	-.51085
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6526.0**	**40939.**	**2061.7**
Stddev	35.4	294.	34.1
%RSD	.54178	.71898	1.6547
#1	6541.9	40985.	2026.5
#2	6550.5	40624.	2063.9
#3	6485.4	41207.	2094.6

Sample Name: LCSW 11 Acquired: 6/23/2016 22:38:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20572	5.1016	.20434	.99099	.52273	.02551	5.4134	.02557
Stddev	.00209	.0119	.00205	.00539	.00783	.00018	.0860	.00030
%RSD	1.0158	.23413	1.0041	.54353	1.4983	.70326	1.5893	1.1859
#1	.20812	5.1093	.20580	.99664	.52935	.02536	5.5127	.02591
#2	.20432	5.1075	.20521	.99043	.52477	.02571	5.3662	.02536
#3	.20472	5.0878	.20199	.98591	.51409	.02545	5.3613	.02542
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10385	.25890	.26189	2.1195	25.823	.53451	5.3645	.26403
Stddev	.00010	.00080	.00098	.0266	.322	.00518	.0615	.00364
%RSD	.09428	.30942	.37489	1.2540	1.2452	.96918	1.1457	1.3773
#1	.10374	.25918	.26254	2.1199	26.125	.53986	5.3994	.26710
#2	.10385	.25953	.26237	2.1460	25.858	.53416	5.2935	.26002
#3	.10394	.25800	.26076	2.0928	25.485	.52952	5.4005	.26498
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.53004	25.957	.26029	4.9232	.25825	.61920	.19466	2.6294
Stddev	.00026	.317	.00111	.0037	.00566	.00409	.00541	.0065
%RSD	.04851	1.2227	.42516	.07504	2.1932	.66079	2.7789	.24514
#1	.53022	26.268	.26156	4.9218	.25403	.62385	.18855	2.6314
#2	.53016	25.970	.25951	4.9204	.25605	.61617	.19656	2.6346
#3	.52975	25.634	.25981	4.9274	.26469	.61757	.19886	2.6222
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: LCSW 11 Acquired: 6/23/2016 22:38:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51706**	**.53317**	**.52168**	**.25600**	**.51571**	**.51503**	**1.3110**
Stddev	.00229	.00553	.01335	.00456	.00094	.00046	.2154
%RSD	.44200	1.0369	2.5594	1.7813	.18322	.08969	16.426
#1	.51564	.53801	.52874	.25386	.51673	.51503	1.4161
#2	.51585	.53436	.53002	.26124	.51555	.51550	1.4536
#3	.51970	.52714	.50628	.25291	.51486	.51457	1.0633
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6527.2**	**40455.**	**2087.1**
Stddev	31.3	163.	12.5
%RSD	.47880	.40351	.59802
#1	6563.2	40283.	2074.8
#2	6508.4	40474.	2086.8
#3	6509.8	40608.	2099.8

Sample Name: F BLANK Acquired: 6/23/2016 22:42:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573353-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00101**	**.01349**	**.00278**	**.00644**	**-.00048**	**.00013**	**.02829**	**-.00023**
Stddev	.00147	.00646	.00238	.00275	.00027	.00015	.01012	.00024
%RSD	145.82	47.931	85.631	42.748	55.833	119.23	35.758	102.61
#1	-.00010	.00837	.00376	.00945	-.00079	.00030	.03838	-.00026
#2	.00267	.02075	.00451	.00583	-.00040	.00003	.02835	-.00045
#3	.00045	.01134	.00007	.00405	-.00027	.00005	.01815	.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00017**	**-.00021**	**.00115**	**.01618**	**-.02636**	**.00022**	**-.03112**	**.00301**
Stddev	.00025	.00041	.00050	.01401	.02024	.00227	.10093	.00056
%RSD	146.69	193.40	43.311	86.581	76.800	1041.8	324.28	18.733
#1	.00022	.00012	.00058	.01388	-.01342	.00194	.07257	.00362
#2	.00039	-.00067	.00140	.00346	-.01597	.00107	-.03691	.00250
#3	-.00010	-.00008	.00148	.03119	-.04969	-.00236	-.12903	.00293
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00043**	**131.96**	**-.00024**	**-.00817**	**-.00215**	**.00148**	**-.00969**	**-.00282**
Stddev	.00010	.85	.00144	.01034	.00162	.00392	.00917	.00327
%RSD	22.241	.64765	607.00	126.51	75.376	266.02	94.572	115.83
#1	.00044	132.92	.00143	.00262	-.00398	.00318	-.01589	.00093
#2	.00053	131.28	-.00110	-.01799	-.00092	.00426	.00084	-.00431
#3	.00034	131.69	-.00104	-.00915	-.00154	-.00301	-.01402	-.00509
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: F BLANK Acquired: 6/23/2016 22:42:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573353-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00016**	**.00010**	**-.00200**	**-.00314**	**.00056**	**.00410**	**.11736**
Stddev	.00046	.00052	.00261	.00021	.00036	.00009	.48090
%RSD	288.96	535.97	130.24	6.7340	64.659	2.2677	409.76
#1	-.00023	.00063	-.00290	-.00292	.00037	.00404	.00662
#2	.00004	.00007	-.00404	-.00316	.00033	.00421	-.29851
#3	.00067	-.00041	.00094	-.00334	.00097	.00406	.64398
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6547.2**	**40321.**	**2096.8**
Stddev	27.3	101.	33.4
%RSD	.41693	.25042	1.5916
#1	6550.7	40326.	2062.9
#2	6572.6	40217.	2129.6
#3	6518.3	40419.	2098.0

Sample Name: F BLANK Acquired: 6/23/2016 22:46:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573353-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00205**	**.01458**	**.00104**	**.00093**	**-.00012**	**-.00011**	**.03413**	**.00007**
Stddev	.00257	.00695	.00149	.00158	.00097	.00015	.06625	.00024
%RSD	125.57	47.660	142.55	168.73	807.82	139.24	194.10	358.50
#1	.00306	.01748	.00276	.00220	-.00059	.00003	.00736	.00006
#2	.00395	.00665	.00032	-.00083	-.00077	-.00008	-.01455	-.00017
#3	-.00087	.01960	.00006	.00144	.00099	-.00027	.10958	.00031
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**-.00025**	**.00229**	**.01454**	**.10361**	**-.00161**	**-.00074**	**.00159**
Stddev	.00020	.00099	.00086	.02039	.12854	.00327	.04972	.00150
%RSD	270.27	392.29	37.466	140.25	124.05	203.19	6704.2	94.513
#1	-.00016	-.00139	.00199	-.00883	.15515	.00182	-.05789	.00331
#2	.00020	.00021	.00326	.02371	.19839	-.00196	.02308	.00082
#3	.00019	.00042	.00163	.02874	-.04269	-.00469	.03259	.00062
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00013**	**.08993**	**.00021**	**-.00044**	**.00171**	**.00115**	**-.00478**	**-.01618**
Stddev	.00021	.02749	.00122	.00310	.00251	.00690	.00614	.00197
%RSD	158.62	30.569	571.20	708.88	146.77	601.02	128.46	12.198
#1	.00007	.10546	-.00067	.00234	-.00022	.00773	-.00669	-.01496
#2	.00036	.10614	-.00029	.00012	.00081	.00175	.00209	-.01845
#3	-.00004	.05819	.00160	-.00377	.00454	-.00604	-.00974	-.01512
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: F BLANK Acquired: 6/23/2016 22:46:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573353-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	.00000	-.00849	-.00355	.00002	.00375	.18786
Stddev	.00024	.00006	.00318	.00112	.00121	.00009	.14076
%RSD	171.57	4326.9	37.397	31.641	7603.4	2.4757	74.927
#1	.00012	.00006	-.01213	-.00271	-.00043	.00385	.34927
#2	.00039	-.00000	-.00626	-.00312	-.00090	.00367	.12376
#3	-.00009	-.00006	-.00708	-.00483	.00138	.00373	.09057
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6645.1	41415.	2118.1
Stddev	51.0	148.	9.8
%RSD	.76790	.35755	.46386
#1	6671.8	41525.	2108.0
#2	6677.3	41246.	2127.6
#3	6586.3	41473.	2118.7

Sample Name: L1606097401 Acquired: 6/23/2016 22:50:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00187	.11141	.00185	.13976	.13124	.00006	11.923	.00025
Stddev	.00099	.00961	.00433	.00304	.00120	.00013	.085	.00026
%RSD	52.908	8.6272	234.23	2.1718	.91402	206.81	.71186	103.09
#1	-.00300	.11755	.00322	.14250	.13178	.00021	12.021	.00001
#2	-.00137	.11635	-.00300	.14028	.13207	-.00003	11.884	.00021
#3	-.00123	.10033	.00533	.13649	.12986	.00001	11.865	.00052
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00325	.00082	.00243	1.0194	3.4016	.00412	14.829	.47953
Stddev	.00065	.00120	.00106	.0300	.0185	.00138	.053	.00284
%RSD	19.954	146.37	43.821	2.9431	.54459	33.491	.35710	.59200
#1	.00322	-.00007	.00334	.98836	3.3804	.00531	14.788	.47741
#2	.00262	.00219	.00268	1.0217	3.4096	.00444	14.889	.48276
#3	.00391	.00035	.00126	1.0482	3.4148	.00261	14.810	.47842
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00016	14.029	.00345	-.00115	.00096	.00344	.00640	6.1973
Stddev	.00012	.046	.00111	.00836	.00159	.00074	.00619	.0079
%RSD	73.635	.32793	32.058	726.14	165.09	21.404	96.675	.12766
#1	-.00003	14.048	.00367	-.00743	.00071	.00392	.01206	6.2055
#2	-.00020	14.062	.00443	-.00436	-.00049	.00379	-.00020	6.1968
#3	-.00025	13.976	.00225	.00834	.00266	.00259	.00734	6.1897
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606097401 Acquired: 6/23/2016 22:50:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00010	.15663	-.00490	-.00350	-.00004	.00381	.27764
Stddev	.00040	.00118	.00403	.00013	.00036	.00005	.25173
%RSD	389.62	.75413	82.244	3.6390	976.32	1.2951	90.667
#1	-.00036	.15760	-.00950	-.00338	-.00044	.00385	.40246
#2	.00031	.15698	-.00318	-.00349	.00013	.00382	.44257
#3	.00036	.15532	-.00201	-.00363	.00021	.00375	-.01211
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6619.1	40945.	2109.7
Stddev	10.2	451.	19.0
%RSD	.15337	1.1019	.89902
#1	6615.2	40453.	2097.8
#2	6611.5	41044.	2131.5
#3	6630.6	41339.	2099.6

Sample Name: L1606097409S Acquired: 6/23/2016 22:54:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20627	5.0015	.19707	1.0840	.63830	.02511	15.216	.02489
Stddev	.00439	.0470	.00646	.0138	.00496	.00023	.159	.00037
%RSD	2.1297	.93911	3.2784	1.2764	.77719	.91601	1.0457	1.4914
#1	.20993	5.0515	.18974	1.0963	.64340	.02535	15.397	.02529
#2	.20140	4.9583	.19954	1.0690	.63801	.02510	15.151	.02456
#3	.20748	4.9946	.20193	1.0866	.63349	.02489	15.099	.02483
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10388	.25321	.25314	2.5849	28.688	.52179	17.865	.68847
Stddev	.00035	.00232	.00112	.0115	.394	.00543	.201	.00731
%RSD	.33918	.91513	.44229	.44279	1.3724	1.0405	1.1266	1.0613
#1	.10359	.25588	.25333	2.5862	29.131	.52794	18.088	.69576
#2	.10427	.25175	.25416	2.5957	28.557	.51978	17.697	.68851
#3	.10377	.25199	.25194	2.5729	28.377	.51766	17.810	.68114
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.51876	38.405	.25390	4.8710	.24934	.59995	.18589	8.2835
Stddev	.00146	.259	.00011	.0108	.00345	.00191	.00207	.0089
%RSD	.28048	.67387	.04392	.22193	1.3854	.31868	1.1116	.10799
#1	.51718	38.681	.25401	4.8591	.24740	.60185	.18570	8.2732
#2	.51906	38.367	.25379	4.8738	.25333	.59803	.18393	8.2891
#3	.52005	38.168	.25392	4.8802	.24730	.59997	.18804	8.2882
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606097409S Acquired: 6/23/2016 22:54:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50349	.65950	.51762	.24758	.50743	.50246	.81618
Stddev	.00174	.00452	.01522	.00062	.00357	.00056	.23521
%RSD	.34477	.68510	2.9397	.25174	.70344	.11237	28.818
#1	.50344	.66471	.52198	.24830	.51109	.50181	.58167
#2	.50178	.65703	.53018	.24729	.50396	.50287	.81481
#3	.50525	.65675	.50070	.24716	.50723	.50268	1.0521
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6571.1	40649.	2099.3
Stddev	19.5	291.	11.1
%RSD	.29604	.71659	.53063
#1	6548.6	40317.	2086.9
#2	6582.6	40767.	2102.9
#3	6582.0	40862.	2108.3

Sample Name: L1606097411SD Acquired: 6/23/2016 22:58:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20301	5.0935	.20254	1.1135	.64720	.02513	17.005	.02529
Stddev	.00131	.0093	.00255	.0073	.00193	.00014	.125	.00014
%RSD	.64434	.18208	1.2591	.65223	.29801	.54856	.73583	.55871
#1	.20180	5.0830	.20062	1.1056	.64841	.02524	16.944	.02527
#2	.20284	5.0968	.20544	1.1153	.64821	.02498	17.149	.02543
#3	.20440	5.1006	.20156	1.1198	.64498	.02518	16.923	.02515
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10473	.25680	.25656	3.0216	28.787	.52889	19.740	.73225
Stddev	.00029	.00169	.00156	.0240	.291	.00510	.166	.00214
%RSD	.27243	.65998	.60790	.79299	1.0111	.96496	.83971	.29254
#1	.10453	.25746	.25836	3.0492	28.796	.52755	19.834	.73428
#2	.10462	.25806	.25556	3.0077	29.074	.53453	19.837	.73246
#3	.10506	.25487	.25576	3.0078	28.492	.52458	19.548	.73001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.52618	39.581	.25576	4.8994	.25448	.60988	.19462	8.7250
Stddev	.00060	.184	.00118	.0146	.00431	.00152	.00488	.0127
%RSD	.11470	.46575	.46133	.29829	1.6939	.24983	2.5089	.14588
#1	.52623	39.585	.25707	4.8974	.25909	.60949	.19392	8.7122
#2	.52676	39.763	.25543	4.9148	.25381	.60859	.19982	8.7250
#3	.52555	39.395	.25478	4.8858	.25055	.61156	.19013	8.7376
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606097411SD Acquired: 6/23/2016 22:58:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50655	.68066	.51426	.25122	.50847	.50678	.50917
Stddev	.00036	.00376	.00731	.00121	.00054	.00037	.18705
%RSD	.07012	.55173	1.4206	.48320	.10539	.07293	36.736
#1	.50668	.67977	.51313	.25144	.50835	.50698	.47043
#2	.50615	.68478	.52206	.25231	.50800	.50700	.34452
#3	.50682	.67742	.50758	.24991	.50905	.50635	.71255
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6599.6	40879.	2125.2
Stddev	36.5	163.	36.7
%RSD	.55319	.39917	1.7271
#1	6558.7	40764.	2124.1
#2	6628.9	40808.	2089.1
#3	6611.1	41066.	2162.5

Sample Name: L1606090301 Acquired: 6/23/2016 23:01:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00171	.01564	.00294	.01850	.00898	-.00004	46.441	.00022
Stddev	.00082	.01258	.00166	.00390	.00087	.00003	.145	.00037
%RSD	47.964	80.448	56.393	21.081	9.6500	63.411	.31216	167.83
#1	.00264	.02660	.00383	.02298	.00987	-.00008	46.595	-.00020
#2	.00144	.01843	.00103	.01664	.00813	-.00002	46.423	.00039
#3	.00106	.00190	.00395	.01587	.00895	-.00004	46.307	.00049
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00038	.00043	.00251	.04392	.36940	-.00187	6.4848	.00401
Stddev	.00055	.00082	.00124	.02745	.02893	.00702	.0718	.00293
%RSD	143.61	190.56	49.243	62.491	7.8320	374.64	1.1071	73.244
#1	.00082	-.00015	.00183	.02878	.35514	-.00147	6.5673	.00719
#2	.00055	.00137	.00394	.07561	.35036	.00493	6.4364	.00342
#3	-.00023	.00007	.00177	.02738	.40269	-.00908	6.4507	.00141
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00124	2.3912	.00015	-.00111	.00175	-.00069	-.00611	2.4464
Stddev	.00019	.0146	.00065	.00585	.00193	.00258	.00753	.0043
%RSD	15.055	.61183	443.31	524.60	110.31	375.66	123.28	.17481
#1	.00123	2.3744	.00089	-.00577	.00397	.00193	-.00911	2.4476
#2	.00106	2.4012	-.00016	.00545	.00055	-.00077	.00246	2.4417
#3	.00143	2.3980	-.00029	-.00303	.00072	-.00322	-.01167	2.4500
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090301 Acquired: 6/23/2016 23:01:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00033	.05094	-.00942	-.00137	.00033	.00855	.12749
Stddev	.00120	.00036	.00450	.00125	.00064	.00008	.22777
%RSD	359.04	.70893	47.793	91.344	191.76	.98957	178.66
#1	-.00095	.05113	-.01374	-.00053	.00015	.00845	.36891
#2	.00142	.05116	-.00475	-.00281	-.00019	.00858	.09713
#3	.00053	.05052	-.00978	-.00077	.00105	.00861	-.08358
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6624.9	41305.	2113.7
Stddev	28.0	611.	12.8
%RSD	.42196	1.4799	.60649
#1	6599.0	41779.	2107.8
#2	6654.5	40615.	2128.4
#3	6621.1	41521.	2104.9

Sample Name: L1606090301PS Acquired: 6/23/2016 23:05:49 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573749-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20315	5.0351	.19933	.98509	.51723	.02533	46.108	.02517
Stddev	.00237	.0141	.00064	.00471	.00225	.00014	.292	.00010
%RSD	1.1667	.28092	.32345	.47796	.43513	.56006	.63289	.38937
#1	.20113	5.0207	.19918	.98035	.51851	.02535	46.344	.02507
#2	.20576	5.0490	.19878	.98516	.51463	.02547	45.781	.02527
#3	.20255	5.0357	.20004	.98977	.51856	.02518	46.198	.02516
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10028	.25408	.25247	2.0837	25.422	.52185	10.797	.26020
Stddev	.00048	.00013	.00281	.0115	.145	.00302	.049	.00165
%RSD	.47862	.05151	1.1142	.55075	.56995	.57813	.44947	.63512
#1	.10024	.25422	.24943	2.0968	25.589	.52416	10.827	.25832
#2	.10077	.25406	.25499	2.0788	25.328	.51844	10.822	.26141
#3	.09982	.25396	.25300	2.0754	25.349	.52296	10.741	.26088
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.52036	27.383	.25017	4.9546	.25694	.60707	.19089	4.8076
Stddev	.00073	.158	.00061	.0081	.00347	.00719	.00485	.0007
%RSD	.13971	.57670	.24497	.16395	1.3493	1.1850	2.5386	.01435
#1	.52119	27.447	.24958	4.9632	.25410	.61287	.18545	4.8079
#2	.51984	27.204	.25012	4.9471	.25592	.60931	.19246	4.8068
#3	.52004	27.500	.25080	4.9535	.26081	.59902	.19476	4.8080
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090301PS Acquired: 6/23/2016 23:05:49 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573749-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50644	.56158	.50038	.25206	.50977	.50584	.50366
Stddev	.00169	.00120	.00839	.00343	.00237	.00060	.30226
%RSD	.33432	.21309	1.6764	1.3594	.46441	.11817	60.013
#1	.50760	.56240	.50354	.25552	.50926	.50626	.73300
#2	.50722	.56020	.49087	.24866	.51235	.50516	.61683
#3	.50450	.56212	.50672	.25200	.50770	.50611	.16114
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6634.2	40772.	2114.5
Stddev	22.4	92.	10.2
%RSD	.33771	.22603	.48307
#1	6611.3	40856.	2104.9
#2	6656.0	40786.	2113.4
#3	6635.2	40673.	2125.2

Sample Name: L1606090301SDL Acquired: 6/23/2016 23:09:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573749-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00125**	**-.00211**	**.00104**	**.00614**	**.00089**	**-.00005**	**8.6724**	**-.00005**
Stddev	.00276	.00609	.00315	.00512	.00101	.00009	.0375	.00023
%RSD	221.14	289.17	302.21	83.373	113.85	164.72	.43247	430.30
#1	.00165	-.00654	-.00255	.00543	.00189	-.00005	8.6376	.00022
#2	.00379	.00484	.00330	.01158	.00093	-.00014	8.6675	-.00019
#3	-.00169	-.00462	.00238	.00141	-.00014	.00003	8.7121	-.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00007**	**-.00015**	**.00151**	**.01774**	**.05719**	**-.00032**	**1.2928**	**.00199**
Stddev	.00032	.00036	.00174	.01658	.07315	.00592	.1368	.00072
%RSD	433.89	237.55	115.25	93.432	127.89	1849.8	10.583	36.311
#1	-.00025	-.00023	.00291	.03261	.10693	.00122	1.1631	.00203
#2	.00038	-.00046	-.00044	-.00013	-.02679	.00468	1.4358	.00125
#3	.00008	.00024	.00207	.02075	.09145	-.00685	1.2797	.00269
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00062**	**.50288**	**.00020**	**.00422**	**-.00347**	**-.00009**	**-.00488**	**.45851**
Stddev	.00015	.01796	.00107	.00310	.00056	.00216	.00737	.00307
%RSD	24.656	3.5717	530.84	73.548	16.155	2299.3	151.12	.67016
#1	.00074	.48937	-.00102	.00065	-.00411	-.00250	-.01008	.45511
#2	.00045	.49600	.00097	.00624	-.00312	.00057	.00356	.46107
#3	.00067	.52326	.00066	.00577	-.00317	.00165	-.00811	.45936
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090301SDL Acquired: 6/23/2016 23:09:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573749-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00043**	**.00975**	**-.00339**	**.00072**	**-.00071**	**.00182**	**.28982**
Stddev	.00076	.00047	.00412	.00137	.00059	.00016	.40983
%RSD	176.59	4.8404	121.60	190.63	83.639	8.7152	141.41
#1	.00038	.00960	.00073	.00044	-.00028	.00190	.42557
#2	-.00111	.01028	-.00752	.00221	-.00047	.00164	.61455
#3	-.00055	.00937	-.00339	-.00049	-.00139	.00193	-.17066
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6811.4**	**42010.**	**2145.9**
Stddev	35.6	211.	19.5
%RSD	.52290	.50301	.90902
#1	6804.9	42193.	2150.1
#2	6779.4	42057.	2163.0
#3	6849.8	41779.	2124.7

Sample Name: CCV Acquired: 6/23/2016 23:13:27 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41326**	**10.290**	**.40933**	**.51318**	**1.0406**	**.05196**	**10.449**
Stddev	.00367	.031	.00303	.00454	.0017	.00056	.073
%RSD	.88729	.29649	.74071	.88463	.16560	1.0752	.70056
#1	.41222	10.266	.40583	.51598	1.0426	.05202	10.433
#2	.41023	10.281	.41097	.50794	1.0395	.05138	10.384
#3	.41734	10.325	.41118	.51561	1.0396	.05249	10.528
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05188**	**.20751**	**.51860**	**.52298**	**4.1404**	**51.810**	**1.0557**
Stddev	.00032	.00026	.00373	.00169	.0293	.081	.0044
%RSD	.62111	.12412	.72010	.32358	.70759	.15585	.41608
#1	.05162	.20739	.51967	.52267	4.1341	51.858	1.0535
#2	.05178	.20781	.51444	.52146	4.1148	51.717	1.0529
#3	.05224	.20734	.52168	.52480	4.1724	51.855	1.0608
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**10.434**	**.52450**	**1.0396**	**51.672**	**.51702**	**10.257**	**.51943**
Stddev	.147	.00403	.0017	.210	.00153	.061	.00236
%RSD	1.4125	.76876	.16574	.40642	.29546	.59705	.45451
#1	10.599	.52887	1.0415	51.851	.51530	10.207	.52009
#2	10.315	.52369	1.0381	51.441	.51819	10.240	.51680
#3	10.387	.52093	1.0392	51.724	.51759	10.325	.52138
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 23:13:27 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2432**	**.40702**	**5.1978**	**1.0337**	**1.0394**	**1.0354**	**.52041**
Stddev	.0039	.00046	.0143	.0031	.0042	.0035	.00214
%RSD	.31258	.11246	.27417	.30049	.40463	.33586	.41209
#1	1.2391	.40753	5.1883	1.0302	1.0436	1.0350	.52283
#2	1.2437	.40689	5.1909	1.0348	1.0352	1.0322	.51875
#3	1.2468	.40665	5.2142	1.0361	1.0396	1.0391	.51965
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0351**	**1.0403**	F **.25266**
Stddev	.0070	.0016	.10822
%RSD	.68141	.15403	42.834
#1	1.0383	1.0386	.34911
#2	1.0270	1.0404	.13562
#3	1.0400	1.0418	.27323
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6474.0**	**39804.**	**2056.5**
Stddev	41.9	151.	9.5
%RSD	.64647	.37833	.46144
#1	6447.9	39827.	2053.8
#2	6522.3	39942.	2048.6
#3	6451.9	39644.	2067.0

Sample Name: CCB Acquired: 6/23/2016 23:17:05 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00053**	**.01688**	**.00315**	**.00457**	**-.00011**	**.00013**	**.01891**	**.00020**
Stddev	.00350	.01032	.00748	.00212	.00110	.00006	.06314	.00005
%RSD	664.12	61.108	237.21	46.535	973.24	44.359	333.92	24.066
#1	-.00268	.02858	-.00372	.00470	.00043	.00018	.03807	.00024
#2	-.00001	.00908	.00206	.00662	.00061	.00007	-.05159	.00021
#3	.00427	.01299	.01112	.00238	-.00138	.00016	.07025	.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00005**	**.00073**	**.00146**	**.01871**	**.09919**	**.00074**	**.03193**	**.00172**
Stddev	.00032	.00177	.00126	.01232	.12450	.00566	.07256	.00242
%RSD	651.38	241.69	86.427	65.838	125.51	769.76	227.22	140.80
#1	.00042	-.00122	.00034	.00498	-.03300	-.00467	.10382	.00139
#2	-.00019	.00221	.00282	.02879	.11637	.00026	-.04127	-.00052
#3	-.00008	.00121	.00121	.02235	.21421	.00662	.03325	.00429
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00396**	**.05479**	**-.00105**	**.00648**	**-.00101**	**.00564**	**-.00498**	**.00630**
Stddev	.00092	.03837	.00180	.00453	.00307	.00402	.00631	.00043
%RSD	23.350	70.024	171.44	69.981	304.27	71.194	126.78	6.8811
#1	.00290	.07487	.00083	.01030	.00189	.00526	-.01221	.00676
#2	.00456	.01055	-.00276	.00766	-.00422	.00984	-.00057	.00622
#3	.00443	.07894	-.00122	.00147	-.00069	.00183	-.00216	.00591
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: CCB Acquired: 6/23/2016 23:17:05 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00055	-.00002	.00408	-.00193	.00037	.00005	.02228
Stddev	.00023	.00019	.00879	.00186	.00041	.00007	.40623
%RSD	41.382	1063.2	215.26	96.532	110.70	135.03	1823.1
#1	.00069	-.00023	-.00341	-.00071	-.00006	-.00002	-.40274
#2	.00029	.00004	.01375	-.00101	.00043	.00013	.06292
#3	.00066	.00013	.00190	-.00407	.00076	.00005	.40667
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6464.5	39816.	2046.0
Stddev	26.9	481.	41.8
%RSD	.41677	1.2077	2.0425
#1	6458.5	39261.	2010.7
#2	6441.1	40106.	2035.2
#3	6493.9	40080.	2092.2

Sample Name: L1606090302 Acquired: 6/23/2016 23:21:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00031	.00800	.00182	.01230	.01036	-.00005	47.673	.00026
Stddev	.00085	.00343	.00363	.00358	.00124	.00007	.280	.00015
%RSD	272.61	42.881	200.18	29.113	11.960	128.20	.58740	57.099
#1	.00031	.00879	-.00237	.00886	.01157	-.00001	47.890	.00037
#2	-.00053	.01096	.00415	.01601	.00909	-.00001	47.772	.00032
#3	.00116	.00424	.00367	.01205	.01042	-.00013	47.357	.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00005	.00015	.00261	.09179	.46952	-.00272	6.7874	.01530
Stddev	.00029	.00088	.00094	.00788	.11796	.00383	.0475	.00163
%RSD	527.39	592.21	36.175	8.5807	25.123	140.65	.69987	10.685
#1	-.00038	.00116	.00159	.08468	.47413	-.00254	6.7436	.01564
#2	.00012	-.00040	.00278	.09044	.34932	.00101	6.7806	.01352
#3	.00010	-.00032	.00345	.10026	.58509	-.00664	6.8379	.01673
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00092	2.5218	-.00025	.00306	.00002	-.00071	-.00373	2.5222
Stddev	.00063	.0302	.00136	.00470	.00238	.00277	.00475	.0109
%RSD	68.779	1.1957	538.26	153.23	11843.	388.81	127.31	.43311
#1	.00032	2.5566	.00100	-.00086	.00039	.00133	.00175	2.5345
#2	.00086	2.5031	-.00169	.00178	.00219	.00040	-.00620	2.5136
#3	.00157	2.5056	-.00007	.00827	-.00252	-.00386	-.00675	2.5186
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090302 Acquired: 6/23/2016 23:21:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00028**	**.05280**	**-.00935**	**-.00377**	**-.00005**	**.00235**	**.11164**
Stddev	.00108	.00027	.00453	.00048	.00146	.00015	.26019
%RSD	392.25	.51917	48.460	12.830	2892.9	6.5370	233.05
#1	-.00049	.05310	-.00574	-.00424	.00086	.00248	.33955
#2	.00090	.05256	-.01444	-.00328	.00072	.00218	-.17184
#3	-.00123	.05274	-.00787	-.00379	-.00173	.00238	.16722
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6545.6**	**41303.**	**2131.1**
Stddev	23.7	132.	13.6
%RSD	.36137	.32048	.63580
#1	6530.1	41250.	2117.0
#2	6572.8	41206.	2132.3
#3	6534.0	41454.	2144.0

Sample Name: L1606090303 Acquired: 6/23/2016 23:25:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00263	.00774	.00563	.00796	.00854	-.00002	45.954	.00013
Stddev	.00328	.00668	.00265	.00318	.00150	.00012	.228	.00034
%RSD	124.84	86.299	47.034	39.934	17.503	682.86	.49687	267.94
#1	.00506	.01059	.00367	.00860	.00804	-.00005	45.886	.00049
#2	-.00111	.01253	.00457	.01077	.01022	.00012	46.208	.00008
#3	.00394	.00011	.00864	.00451	.00737	-.00012	45.767	-.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00022	-.00004	.00264	.21996	.44485	.00448	6.7971	.02189
Stddev	.00017	.00063	.00127	.02132	.06435	.00051	.0747	.00331
%RSD	77.701	1729.3	48.231	9.6943	14.465	11.277	1.0984	15.131
#1	.00007	.00062	.00404	.23944	.50211	.00394	6.7547	.01807
#2	.00019	-.00009	.00157	.19718	.45723	.00494	6.8833	.02378
#3	.00041	-.00064	.00230	.22325	.37521	.00457	6.7534	.02384
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00090	2.4891	-.00031	.00226	-.00318	.00121	-.00101	2.4462
Stddev	.00023	.0385	.00053	.00868	.00252	.00105	.00896	.0110
%RSD	25.268	1.5452	169.73	384.84	79.079	87.283	885.01	.44916
#1	.00089	2.4503	-.00059	-.00015	-.00309	-.00001	-.00785	2.4500
#2	.00067	2.4900	.00030	.01188	-.00575	.00184	-.00433	2.4549
#3	.00113	2.5272	-.00064	-.00497	-.00072	.00180	.00914	2.4339
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090303 Acquired: 6/23/2016 23:25:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.05042	-.00867	-.00405	-.00058	.00293	.21469
Stddev	.00068	.00039	.00321	.00395	.00086	.00012	.24147
%RSD	116.53	.76868	37.044	97.556	147.98	4.1433	112.48
#1	.00096	.05045	-.00617	-.00495	-.00048	.00287	.22391
#2	.00100	.05080	-.01229	-.00746	-.00148	.00307	.45141
#3	-.00020	.05002	-.00756	.00027	.00022	.00286	-.03127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6557.1	40614.	2090.8
Stddev	3.1	286.	4.7
%RSD	.04661	.70434	.22607
#1	6559.9	40616.	2093.3
#2	6553.9	40327.	2085.3
#3	6557.7	40899.	2093.7

Sample Name: L1606090304 Acquired: 6/23/2016 23:28:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00249**	**.01012**	**.00464**	**.01066**	**.00806**	**-.00002**	**46.697**	**.00023**
Stddev	.00091	.01191	.00138	.00192	.00037	.00002	.216	.00013
%RSD	36.530	117.71	29.729	17.988	4.5683	108.87	.46335	55.814
#1	.00283	-.00243	.00466	.01137	.00780	-.00002	46.940	.00033
#2	.00319	.01152	.00601	.00849	.00791	.00000	46.527	.00026
#3	.00146	.02127	.00325	.01212	.00849	-.00003	46.623	.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00033**	**.00062**	**.00100**	**.51001**	**.64738**	**.00217**	**6.6353**	**.05303**
Stddev	.00037	.00064	.00122	.03061	.09484	.00222	.1713	.00028
%RSD	109.48	103.02	122.22	6.0022	14.650	102.21	2.5823	.52039
#1	-.00001	.00028	.00000	.50309	.73158	.00219	6.5308	.05285
#2	-.00026	.00023	.00063	.48345	.66592	.00439	6.8331	.05291
#3	-.00073	.00136	.00236	.54349	.54463	-.00006	6.5421	.05335
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00050**	**2.5222**	**.00010**	**.00174**	**.00050**	**.00128**	**-.00532**	**2.7225**
Stddev	.00078	.0513	.00045	.01103	.00453	.00439	.00901	.0008
%RSD	156.95	2.0349	434.03	634.47	913.70	343.72	169.31	.03118
#1	-.00033	2.5754	.00043	-.00461	.00243	-.00315	-.01182	2.7217
#2	.00121	2.4730	-.00041	.01448	.00374	.00135	-.00910	2.7234
#3	.00061	2.5183	.00029	-.00465	-.00468	.00562	.00496	2.7225
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090304 Acquired: 6/23/2016 23:28:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00081**	**.04960**	**-.00329**	**-.00196**	**-.00058**	**.00257**	**.19345**
Stddev	.00125	.00047	.00925	.00100	.00022	.00015	.14173
%RSD	153.83	.94383	281.36	50.890	37.305	5.7376	73.263
#1	-.00194	.04918	-.00253	-.00307	-.00034	.00241	.33632
#2	.00053	.04951	.00556	-.00167	-.00074	.00270	.19114
#3	-.00102	.05011	-.01289	-.00114	-.00067	.00259	.05289
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6594.2**	**40470.**	**2074.5**
Stddev	30.7	142.	46.3
%RSD	.46503	.35187	2.2319
#1	6626.4	40634.	2027.0
#2	6565.3	40392.	2077.0
#3	6590.8	40384.	2119.5

Sample Name: L1606090305 Acquired: 6/23/2016 23:32:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00213	.01717	-.00084	.01089	.00981	.00002	48.923	.00033
Stddev	.00269	.00400	.00320	.00082	.00051	.00009	.164	.00010
%RSD	126.20	23.278	380.73	7.5221	5.1474	403.64	.33444	29.068
#1	-.00080	.01450	.00248	.01170	.01029	.00012	49.107	.00031
#2	.00271	.02176	-.00390	.01090	.00928	.00001	48.795	.00044
#3	.00448	.01524	-.00110	.01007	.00986	-.00006	48.866	.00025
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00020	-.00068	.00177	1.5788	.44150	.00018	6.7726	.08864
Stddev	.00038	.00111	.00108	.0172	.06024	.00100	.1054	.00177
%RSD	191.06	162.17	60.718	1.0921	13.644	545.86	1.5565	1.9982
#1	-.00014	-.00131	.00085	1.5835	.49954	.00095	6.8808	.08930
#2	.00013	.00060	.00296	1.5597	.37928	.00054	6.7668	.08663
#3	.00060	-.00134	.00152	1.5932	.44569	-.00095	6.6702	.08998
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00035	2.4907	.00039	.00113	.00257	.00234	-.00366	3.0530
Stddev	.00062	.0095	.00077	.00213	.00167	.00385	.01135	.0045
%RSD	175.61	.38135	200.05	188.45	64.988	164.32	310.04	.14641
#1	-.00035	2.4923	-.00007	.00130	.00338	.00574	-.00469	3.0528
#2	.00078	2.4993	.00128	-.00108	.00370	-.00184	-.01446	3.0486
#3	.00062	2.4805	-.00005	.00317	.00065	.00313	.00817	3.0576
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090305 Acquired: 6/23/2016 23:32:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00063**	**.05089**	**-.00933**	**-.00366**	**-.00080**	**.01192**	**.09534**
Stddev	.00059	.00042	.00867	.00169	.00053	.00018	.35498
%RSD	93.514	.82669	92.914	46.068	66.583	1.5108	372.34
#1	-.00036	.05120	-.01547	-.00172	-.00063	.01174	-.06523
#2	-.00022	.05105	.00059	-.00451	-.00139	.01210	.50224
#3	-.00130	.05041	-.01311	-.00475	-.00037	.01192	-.15099
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6602.9**	**40942.**	**2135.9**
Stddev	6.9	324.	25.7
%RSD	.10425	.79054	1.2024
#1	6597.6	40623.	2114.4
#2	6610.7	41270.	2164.3
#3	6600.5	40933.	2129.0

Sample Name: L1606092401 Acquired: 6/23/2016 23:36:50 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00265	.03744	.60307	2.0253	.01194	.00006	59.280
Stddev	.00277	.00788	.00343	.0045	.00047	.00007	.270
%RSD	104.56	21.060	.56883	.22109	3.9165	122.30	.45465
#1	.00477	.04143	.60414	2.0202	.01227	.00013	59.521
#2	-.00048	.02835	.59924	2.0278	.01216	-.00002	58.989
#3	.00365	.04252	.60585	2.0280	.01141	.00007	59.330
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00037	.00169	.04543	.00223	.02086	20.304	.12271
Stddev	.00012	.00069	.00181	.00172	.03477	.327	.00387
%RSD	33.446	40.936	3.9734	76.929	166.67	1.6108	3.1576
#1	.00045	.00112	.04573	.00336	.05783	20.550	.12442
#2	.00044	.00245	.04349	.00026	.01592	20.429	.12545
#3	.00023	.00148	.04706	.00308	-.01117	19.933	.11828
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.44128	.00138	.09389	F 925.96	-.00083	.00321	-.00198
Stddev	.02883	.00107	.00073	20.38	.00037	.00049	.00073
%RSD	6.5329	77.799	.77502	2.2006	45.088	15.416	37.027
#1	.47443	.00015	.09321	948.64	-.00126	.00268	-.00137
#2	.42732	.00186	.09382	909.19	-.00057	.00328	-.00178
#3	.42208	.00212	.09466	920.06	-.00066	.00367	-.00280
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Sample Name: L1606092401 Acquired: 6/23/2016 23:36:50 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02428**	**.03016**	**5.4351**	**-.00127**	**.13763**	**-.01760**	**-.00342**
Stddev	.00750	.01129	.0055	.00172	.00098	.00201	.00308
%RSD	30.891	37.425	.10042	135.11	.71182	11.421	90.086
#1	.02439	.02726	5.4392	.00004	.13847	-.01734	-.00405
#2	.03173	.04262	5.4371	-.00064	.13656	-.01973	-.00008
#3	.01673	.02061	5.4289	-.00322	.13787	-.01573	-.00615
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.00453**	**.02225**	F **-.12201**
Stddev	.00050	.00015	.18874
%RSD	11.054	.67831	154.69
#1	.00443	.02208	-.21896
#2	.00408	.02238	-.24258
#3	.00507	.02228	.09550
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6296.4**	**38355.**	**2086.6**
Stddev	9.5	124.	22.4
%RSD	.15057	.32201	1.0746
#1	6287.1	38290.	2062.1
#2	6296.1	38498.	2106.2
#3	6306.0	38278.	2091.5

Sample Name: DI7 Acquired: 6/23/2016 23:40:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00122	-.00305	.00325	.01342	.00511	-.00003	.63457	.00006
Stddev	.00070	.00725	.00183	.00147	.00092	.00006	.03174	.00026
%RSD	57.204	237.69	56.132	10.964	18.001	169.74	5.0023	454.53
#1	.00129	-.01124	.00390	.01220	.00527	-.00006	.66854	.00019
#2	.00049	.00254	.00119	.01301	.00412	-.00007	.60566	-.00024
#3	.00189	-.00045	.00467	.01505	.00594	.00003	.62951	.00022
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00034	-.00105	.00137	.04887	.17552	-.00411	.67989	.02289
Stddev	.00029	.00066	.00135	.03758	.14542	.00295	.10949	.00017
%RSD	86.943	62.496	98.260	76.910	82.851	71.791	16.104	.73432
#1	.00007	-.00034	.00163	.08822	.25845	-.00273	.72653	.02297
#2	.00065	-.00118	.00257	.01334	.00761	-.00210	.75834	.02270
#3	.00029	-.00164	-.00009	.04504	.26051	-.00749	.55480	.02300
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00034	.88120	.00036	.00359	-.00382	-.00015	.00024	.27765
Stddev	.00027	.03865	.00120	.00516	.00377	.00137	.00372	.00507
%RSD	80.796	4.3866	328.28	143.80	98.655	910.66	1547.5	1.8265
#1	-.00036	.92576	.00056	.00038	-.00770	-.00061	-.00359	.27282
#2	-.00006	.85668	-.00092	.00954	-.00358	.00139	.00047	.27719
#3	-.00060	.86116	.00145	.00084	-.00017	-.00123	.00384	.28294
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: DI7 Acquired: 6/23/2016 23:40:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00014**	**.00711**	**-.00708**	**-.00329**	**-.00065**	**.00225**	**.10588**
Stddev	.00142	.00039	.01416	.00145	.00012	.00019	.39638
%RSD	991.52	5.5260	199.93	44.028	18.378	8.2791	374.37
#1	-.00086	.00703	-.01555	-.00371	-.00072	.00228	.22011
#2	.00150	.00677	.00926	-.00449	-.00071	.00242	-.33508
#3	-.00107	.00754	-.01496	-.00168	-.00051	.00205	.43261
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6937.5**	**42404.**	**2145.0**
Stddev	31.4	540.	24.8
%RSD	.45241	1.2743	1.1557
#1	6923.1	42362.	2162.5
#2	6915.9	42964.	2155.8
#3	6973.5	41886.	2116.6

Sample Name: DI8 Acquired: 6/23/2016 23:44:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01092	.24637	.01430	.05175	.02804	.00125	.70558	.00137
Stddev	.00255	.00988	.00240	.00085	.00181	.00008	.05906	.00031
%RSD	23.343	4.0097	16.798	1.6422	6.4611	6.1453	8.3704	22.665
#1	.01275	.23892	.01336	.05090	.02859	.00117	.73476	.00163
#2	.01199	.25758	.01250	.05173	.02602	.00133	.63761	.00146
#3	.00801	.24262	.01702	.05260	.02952	.00124	.74437	.00102
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00507	.01122	.01287	.11447	1.4023	.02306	.83494	.03245
Stddev	.00024	.00067	.00189	.02802	.1110	.00136	.03948	.00268
%RSD	4.7987	6.0062	14.707	24.480	7.9171	5.8811	4.7291	8.2701
#1	.00493	.01046	.01351	.14674	1.4178	.02155	.80587	.03520
#2	.00493	.01147	.01074	.10042	1.2843	.02345	.81905	.03233
#3	.00535	.01173	.01436	.09625	1.5047	.02418	.87989	.02983
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02127	1.8562	.01215	.23280	.01067	.03011	.00922	.37328
Stddev	.00053	.0342	.00120	.00909	.00269	.00363	.00276	.00458
%RSD	2.4934	1.8408	9.8651	3.9061	25.214	12.073	29.987	1.2271
#1	.02172	1.8278	.01327	.24205	.00898	.03426	.00781	.37038
#2	.02141	1.8468	.01231	.22387	.01377	.02752	.01240	.37091
#3	.02069	1.8941	.01088	.23248	.00925	.02853	.00745	.37856
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: DI8 Acquired: 6/23/2016 23:44:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02095	.02967	.01993	.00726	.02346	.02475	.30772
Stddev	.00032	.00011	.00361	.00564	.00093	.00004	.15766
%RSD	1.5278	.37363	18.102	77.691	3.9489	.16512	51.234
#1	.02080	.02956	.01706	.01372	.02253	.02472	.37227
#2	.02072	.02978	.02398	.00329	.02439	.02479	.42286
#3	.02131	.02968	.01875	.00477	.02345	.02473	.12803
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6883.6	42542.	2132.1
Stddev	35.9	81.	32.8
%RSD	.52157	.19075	1.5391
#1	6873.8	42453.	2096.2
#2	6923.4	42611.	2160.5
#3	6853.7	42564.	2139.6

Sample Name: DI9 Acquired: 6/23/2016 23:48:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.01172**	**.24362**	**.00993**	**.05401**	**.02815**	**.00127**	**.79730**
Stddev	.00152	.00569	.00166	.00158	.00083	.00011	.02246
%RSD	12.924	2.3335	16.753	2.9338	2.9467	8.5602	2.8175
#1	.01069	.24626	.01067	.05290	.02910	.00116	.79850
#2	.01102	.23710	.01110	.05583	.02755	.00137	.81914
#3	.01346	.24751	.00803	.05330	.02780	.00127	.77426
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00129**	**.00488**	**.01079**	**.01208**	**.12723**	**1.2786**	**.02511**
Stddev	.00009	.00006	.00031	.00168	.01475	.0943	.00185
%RSD	6.9948	1.2164	2.8343	13.930	11.594	7.3784	7.3599
#1	.00122	.00481	.01066	.01339	.14076	1.1744	.02529
#2	.00139	.00493	.01057	.01268	.11150	1.3582	.02686
#3	.00125	.00489	.01114	.01018	.12944	1.3031	.02318
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.89576**	**.03366**	**.02131**	**1.9045**	**.01104**	**.23277**	**.01075**
Stddev	.05721	.00137	.00010	.0400	.00057	.00507	.00231
%RSD	6.3863	4.0674	.47440	2.0992	5.1553	2.1800	21.454
#1	.95963	.03212	.02142	1.8638	.01067	.23321	.01329
#2	.84922	.03409	.02122	1.9059	.01170	.22750	.00878
#3	.87844	.03476	.02130	1.9437	.01076	.23762	.01018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: DI9 Acquired: 6/23/2016 23:48:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.03181**	**.00568**	**.38831**	**.02056**	**.03131**	**.02182**	**.01037**
Stddev	.00137	.01144	.00136	.00010	.00017	.00825	.00275
%RSD	4.3124	201.38	.34995	.48690	.53387	37.784	26.554
#1	.03338	.01685	.38970	.02064	.03120	.02474	.00719
#2	.03087	-.00602	.38824	.02045	.03123	.02822	.01188
#3	.03117	.00622	.38698	.02060	.03150	.01252	.01204
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.02300**	**.02452**	F **-.17091**
Stddev	.00069	.00018	.04679
%RSD	2.9886	.74679	27.377
#1	.02312	.02467	-.13285
#2	.02363	.02457	-.15673
#3	.02227	.02431	-.22315
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6863.0**	**42455.**	**2140.2**
Stddev	39.9	478.	31.6
%RSD	.58151	1.1260	1.4765
#1	6898.2	41962.	2103.8
#2	6819.6	42485.	2156.2
#3	6871.2	42917.	2160.6

Sample Name: CCV Acquired: 6/23/2016 23:52:52 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.41335	10.421	.41508	.51836	1.0496	.05256	10.492
Stddev	.00039	.045	.00377	.00553	.0024	.00034	.011
%RSD	.09500	.43207	.90936	1.0674	.22730	.64627	.10764
#1	.41368	10.369	.41539	.51316	1.0470	.05217	10.480
#2	.41292	10.443	.41116	.51776	1.0502	.05276	10.503
#3	.41347	10.451	.41869	.52417	1.0517	.05276	10.494
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05201	.20974	.52865	.52594	4.2034	52.520	1.0760
Stddev	.00016	.00127	.00164	.00103	.0566	.261	.0077
%RSD	.30575	.60470	.30961	.19638	1.3473	.49635	.71934
#1	.05219	.20874	.52729	.52653	4.2121	52.227	1.0702
#2	.05189	.20932	.52820	.52475	4.1429	52.726	1.0848
#3	.05194	.21117	.53047	.52654	4.2552	52.606	1.0729
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	10.516	.52754	1.0503	52.353	.52349	10.349	.52208
Stddev	.065	.00345	.0031	.035	.00255	.035	.00042
%RSD	.61612	.65474	.29626	.06645	.48722	.33829	.08130
#1	10.587	.52445	1.0535	52.313	.52122	10.327	.52177
#2	10.461	.53127	1.0472	52.379	.52299	10.332	.52191
#3	10.500	.52692	1.0503	52.366	.52625	10.390	.52256
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 23:52:52 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2512**	**.41632**	**5.2448**	**1.0458**	**1.0501**	**1.0483**	**.52498**
Stddev	.0079	.00080	.0298	.0049	.0030	.0126	.00266
%RSD	.62850	.19294	.56895	.46593	.28810	1.2026	.50623
#1	1.2568	.41583	5.2243	1.0406	1.0479	1.0390	.52296
#2	1.2422	.41589	5.2311	1.0465	1.0488	1.0433	.52399
#3	1.2546	.41725	5.2791	1.0503	1.0535	1.0626	.52799
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0496**	**1.0495**	F **.09970**
Stddev	.0053	.0039	.10577
%RSD	.50966	.36953	106.08
#1	1.0434	1.0481	.19784
#2	1.0529	1.0466	.11360
#3	1.0524	1.0539	-.01233
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6372.6**	**39443.**	**2025.2**
Stddev	36.3	113.	32.7
%RSD	.56912	.28682	1.6135
#1	6379.4	39570.	1991.0
#2	6404.9	39404.	2028.4
#3	6333.4	39354.	2056.1

Sample Name: CCB Acquired: 6/23/2016 23:56:30 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00038**	**.01156**	**.00056**	**.00234**	**.00056**	**.00010**	**.03716**
Stddev	.00349	.01000	.00359	.00308	.00040	.00012	.04368
%RSD	913.43	86.498	638.46	132.02	72.010	128.40	117.56
#1	.00278	.00590	.00379	.00161	.00011	-.00003	.06293
#2	.00200	.02311	.00120	.00572	.00089	.00010	-.01328
#3	-.00362	.00567	-.00330	-.00032	.00068	.00022	.06182
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00024**	**.00016**	**.00067**	**.00127**	**.01578**	**.03402**	**-.00346**
Stddev	.00007	.00014	.00147	.00047	.02063	.10839	.00477
%RSD	29.479	83.436	221.45	36.697	130.75	318.63	137.81
#1	.00027	.00008	-.00033	.00181	.00506	.08712	-.00283
#2	.00029	.00009	-.00003	.00096	.00272	-.09068	-.00851
#3	.00016	.00032	.00236	.00106	.03957	.10561	.00096
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00827**	**.00236**	**.00382**	**.04976**	**.00076**	**.00558**	**-.00126**
Stddev	.03186	.00300	.00012	.02808	.00054	.00951	.00325
%RSD	385.40	127.33	3.0647	56.434	70.960	170.62	257.97
#1	-.01385	.00576	.00377	.02300	.00097	.01389	-.00238
#2	-.00614	.00008	.00395	.07899	.00015	-.00480	.00240
#3	.04479	.00123	.00373	.04728	.00116	.00763	-.00379
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 23:56:30 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00469**	**-.00110**	**.00540**	**-.00045**	**.00020**	**-.00579**	**-.00171**
Stddev	.00519	.00566	.00080	.00082	.00023	.00790	.00277
%RSD	110.60	514.68	14.743	182.74	110.97	136.58	162.08
#1	.00345	.00405	.00520	.00050	.00023	-.00735	-.00373
#2	.01039	-.00018	.00628	-.00094	.00042	-.01279	.00145
#3	.00024	-.00717	.00473	-.00090	-.00003	.00278	-.00284
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00035**	**.00012**	F **-.30075**
Stddev	.00143	.00011	.38164
%RSD	413.19	92.936	126.90
#1	.00023	.00022	-.11902
#2	.00071	-.00000	-.73929
#3	-.00198	.00015	-.04394
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6443.5**	**39921.**	**2022.5**
Stddev	30.3	559.	26.8
%RSD	.46979	1.3996	1.3238
#1	6467.0	40295.	2021.5
#2	6409.3	40190.	1996.3
#3	6454.1	39279.	2049.8

Sample Name: ICSA Acquired: 6/24/2016 0:00:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	s -.00181	^F *****	k .00039	s .01077	F -16.124	s .00066
Stddev	.00405	-----	.00084	.01107	27.933	.00100
%RSD	224.07	-----	217.80	102.72	173.24	151.08
#1	s .00084	s .38723	k .00033	s .00228	-48.378	s -.00024
#2	s .00021	s 13.731	k .00125	s .00675	.0154	s .00048
#3	s -.00648	^ -----	k -.00043	s .02329	-.0093	s .00173
Check ?	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit		300.00			.00400	
Low Limit		200.00			-.00400	

Elem	Ca4226	Cd2288	Co2286	Cr2677	Cu2247	Fe2611
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F -17.795	kF -.00145	k .00006	s .00397	kF .00900	kF 13.080
Stddev	30.964	.00234	.00025	.00687	.00196	22.623
%RSD	174.00	161.85	431.98	173.11	21.777	172.95
#1	-53.549	k -.00415	k .00030	s -.00220	k .00711	k 39.203
#2	.1724	k .00000	k .00007	s .00273	k .00885	k .00481
#3	-.0087	k -.00020	k -.00020	s .01137	k .01102	k .03331
Check ?	Chk Fail	Chk Fail	Chk Pass	Chk Pass	Chk Fail	Chk Fail
High Limit	300.00	.00080			.00400	120.00
Low Limit	200.00	-.00080			-.00400	80.000

Elem	K_7664	Li6707	Mg2790	Mn2576	Mo2020	Na5895
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F 303.15	F 17.009	kF 144.99	kF -2.2313	k .00055	F 349.42
Stddev	524.72	29.469	251.19	3.8674	.00057	605.42
%RSD	173.09	173.26	173.25	173.32	104.51	173.26
#1	909.05	51.038	k 435.03	k -6.6970	k -.00009	1048.5
#2	.2954	-.01212	k -.0973	k .00125	k .00102	-.464
#3	.1091	.00231	k .0233	k .00181	k .00071	.231
Check ?	Chk Fail	Chk Fail	Chk Fail	Chk Fail	Chk Pass	Chk Fail
High Limit	.40000	.08000	300.00	.00400		.25000
Low Limit	-.40000	-.08000	200.00	-.00400		-.25000

Sample Name: ICSA Acquired: 6/24/2016 0:00:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	k **.00287**	k **.00325**	kF **.01305**	k **.00093**	k **-.00095**	k **-.02524**
Stddev	.00051	.00655	.00114	.00346	.00637	.00077
%RSD	17.929	201.56	8.7344	372.00	672.86	3.0546
#1	k .00291	k .00302	k .01321	k .00198	k -.00342	k -.02442
#2	k .00336	k .00992	k .01183	k .00374	k .00629	k -.02594
#3	k .00234	k -.00319	k .01409	k -.00293	k -.00571	k -.02536
Check ?	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit			.00400			
Low Limit			-.00400			

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	k **-.00038**	F **-.53580**	kF **27.418**	k **.00263**	s **-.00151**	k **.00080**
Stddev	.00064	.93031	47.540	.00108	.00198	.00005
%RSD	165.52	173.63	173.39	41.177	131.41	6.5158
#1	k -.00073	-1.6100	k 82.313	k .00161	s .00053	k .00074
#2	k -.00077	.00282	k -.07422	k .00377	s -.00163	k .00084
#3	k .00035	-.00020	k .01560	k .00250	s -.00343	k .00082
Check ?	Chk Pass	Chk Fail	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit		.00800	.02400			
Low Limit		-.00800	-.02400			

Elem	Zr3391
Units	ppm
Avg	kF **2038.2**
Stddev	3532.4
%RSD	173.31
#1	k 6117.0
#2	k -4.501
#3	k 2.059
Check ?	Chk Fail
High Limit	.02000
Low Limit	-.02000

Sample Name: ICSA Acquired: 6/24/2016 0:00:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9727.2**	^ *****	**2367.9**
Stddev	176.1	-----	3403.9
%RSD	1.8108	-----	143.75
#1	9699.2	^ -----	-.533
#2	9915.7	^ -----	835.60
#3	9566.8	^ -----	6268.6

Sample Name: ICSAB Acquired: 6/24/2016 0:04:29 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	sF .03510	^F *****	kF .02243	s -.00104	F -17.624	sF .01622
Stddev	.03890	-----	.00215	.00199	39.067	.01920
%RSD	110.83	-----	9.5662	190.62	221.67	118.34
#1	s .00126	s -.05314	k .01995	s -.00051	-62.399	s -.00008
#2	s .02644	^ -----	k .02376	s .00063	9.5264	s .01137
#3	s .07761	s 36.789	k .02358	s -.00324	-.0004	s .03738
Check ?	Chk Fail	Chk Fail	Chk Fail	Chk Pass	Chk Fail	Chk Fail
High Limit	.60000	300.00	.30000		.30000	.30000
Low Limit	.40000	200.00	.20000		.20000	.20000

Elem	Ca4226	Cd2288	Co2286	Cr2677	Cu2247	Fe2611
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F -41.394	kF .04368	kF .02300	sF .01591	kF .03181	kF -23.435
Stddev	80.483	.00655	.00163	.02148	.00204	29.861
%RSD	194.43	14.997	7.0968	135.04	6.3986	127.42
#1	-134.15	k .03613	k .02112	s -.00134	k .02989	k -57.047
#2	9.956	k .04777	k .02379	s .00909	k .03394	k -13.288
#3	.012	k .04715	k .02408	s .03996	k .03158	k .0307
Check ?	Chk Fail	Chk Fail	Chk Fail	Chk Fail	Chk Fail	Chk Fail
High Limit	300.00	.60000	.30000	.30000	.30000	120.00
Low Limit	200.00	.40000	.20000	.20000	.20000	80.000

Elem	K_7664	Li6707	Mg2790	Mn2576	Mo2020	Na5895
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F 64.851	F 12.009	kF -20.121	kF 2.2589	k .00017	F 379.32
Stddev	297.64	30.686	17.446	6.5094	.00008	850.62
%RSD	458.96	255.52	86.705	288.17	49.010	224.25
#1	389.61	46.885	k -30.037	k 9.5964	k .00008	1353.6
#2	-194.93	-10.851	k -30.348	k -2.8212	k .00024	-215.77
#3	-.126	-.0058	k .0229	k .00136	k .00019	.137
Check ?	Chk Fail	Chk Fail	Chk Fail	Chk Fail	Chk Pass	Chk Fail
High Limit	6.0000	.10000	300.00	.30000		6.0000
Low Limit	4.0000	-.10000	200.00	.20000		4.0000

Sample Name: ICSAB Acquired: 6/24/2016 0:04:29 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	kF **.04829**	k **-.00766**	kF **.05850**	kF **.04638**	kF **.02738**	k **-.02638**
Stddev	.00365	.00766	.00306	.00425	.00303	.00499
%RSD	7.5565	99.904	5.2255	9.1610	11.057	18.931
#1	k .04412	k -.00318	k .05693	k .04159	k .03052	k -.02995
#2	k .05085	k -.00330	k .05656	k .04971	k .02448	k -.02852
#3	k .04991	k -.01650	k .06203	k .04783	k .02714	k -.02067
Check ?	Chk Fail	Chk Pass	Chk Fail	Chk Fail	Chk Fail	Chk Pass
High Limit	.60000		.60000	.60000	.30000	
Low Limit	.40000		.40000	.40000	.20000	

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	k **-.00052**	F **.44241**	kF **29.066**	kF **.04366**	sF **.01717**	kF **.04537**
Stddev	.00035	.38366	66.439	.00317	.02029	.00340
%RSD	66.984	86.721	228.58	7.2485	118.19	7.4898
#1	k -.00035	.66466	k 105.09	k .04023	s -.00018	k .04144
#2	k -.00091	.66318	k -17.882	k .04647	s .01220	k .04733
#3	k -.00029	-.00061	k -.0063	k .04429	s .03948	k .04732
Check ?	Chk Pass	Chk Fail	Chk Fail	Chk Fail	Chk Fail	Chk Fail
High Limit		.01000	.03000	.60000	.30000	.60000
Low Limit		-.01000	-.03000	.40000	.20000	.40000

Elem	Zr3391
Units	ppm
Avg	kF **4441.9**
Stddev	9705.7
%RSD	218.50
#1	k 15574.
#2	k -2248.1
#3	k .15
Check ?	Chk Fail
High Limit	.02500
Low Limit	-.02500

Sample Name: ICSAB Acquired: 6/24/2016 0:04:29 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9862.8	^ *****	713.41
Stddev	938.9	-----	1233.6
%RSD	9.5192	-----	172.91
#1	10665.	^ -----	-.471
#2	10093.	^ -----	2.867
#3	8830.2	^ -----	2137.8

Sample Name: CCV Acquired: 6/24/2016 0:08:22 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40571**	**10.252**	**.40176**	**.50352**	**1.0259**	**.05216**	**10.249**
Stddev	.00299	.036	.00436	.00116	.0046	.00015	.046
%RSD	.73722	.35335	1.0850	.22983	.44349	.28067	.44972
#1	.40902	10.294	.39676	.50219	1.0215	.05229	10.217
#2	.40321	10.228	.40475	.50401	1.0306	.05200	10.229
#3	.40489	10.234	.40377	.50434	1.0255	.05219	10.302
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05061**	**.20363**	**.51898**	**.51228**	**4.1370**	**50.997**	**1.0462**
Stddev	.00046	.00031	.00258	.00236	.0140	.262	.0028
%RSD	.90962	.15400	.49724	.46108	.33750	.51330	.27175
#1	.05077	.20360	.52152	.50960	4.1210	50.850	1.0429
#2	.05096	.20396	.51636	.51314	4.1438	51.299	1.0477
#3	.05008	.20334	.51907	.51408	4.1463	50.841	1.0479
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**10.348**	**.51498**	**1.0223**	**50.946**	**.51068**	**10.026**	**.51362**
Stddev	.113	.00344	.0028	.211	.00210	.018	.00135
%RSD	1.0914	.66705	.27479	.41321	.41141	.17759	.26270
#1	10.350	.51110	1.0252	50.805	.51092	10.041	.51220
#2	10.460	.51621	1.0220	51.188	.50847	10.031	.51488
#3	10.234	.51763	1.0196	50.845	.51265	10.006	.51378
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/24/2016 0:08:22 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2074**	**.40666**	**5.0928**	**1.0152**	**1.0270**	**1.0159**	**.51326**
Stddev	.0053	.00225	.0089	.0020	.0057	.0060	.00130
%RSD	.43925	.55339	.17475	.20078	.55230	.59351	.25324
#1	1.2134	.40912	5.0829	1.0130	1.0205	1.0208	.51297
#2	1.2053	.40614	5.1001	1.0158	1.0309	1.0178	.51469
#3	1.2034	.40471	5.0956	1.0169	1.0297	1.0092	.51214
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0354**	**1.0194**	F **.15885**
Stddev	.0015	.0011	.28455
%RSD	.14387	.10553	179.12
#1	1.0371	1.0182	.48730
#2	1.0345	1.0199	-.01296
#3	1.0347	1.0202	.00222
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6365.1**	**39110.**	**2005.2**
Stddev	24.8	486.	19.1
%RSD	.39033	1.2423	.95151
#1	6367.8	38552.	2019.3
#2	6388.5	39438.	1983.5
#3	6339.0	39340.	2012.9

Sample Name: CCB Acquired: 6/24/2016 0:12:00 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00050**	**.01206**	**.00160**	**.00292**	**-.00076**	**.00002**	**.02662**	**.00027**
Stddev	.00084	.00666	.00259	.00138	.00101	.00006	.04195	.00026
%RSD	168.40	55.161	162.13	47.221	133.70	271.30	157.60	96.736
#1	-.00086	.01566	-.00084	.00208	-.00074	.00001	-.00269	.00055
#2	-.00111	.01615	.00432	.00217	.00024	.00009	.07467	.00003
#3	.00046	.00439	.00131	.00451	-.00178	-.00003	.00787	.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00031**	**-.00012**	**.00191**	**.00930**	**-.00712**	**-.00214**	**.08690**	**.00263**
Stddev	.00007	.00094	.00113	.01540	.09612	.00388	.02902	.00322
%RSD	21.273	789.62	59.288	165.71	1349.4	181.57	33.396	122.26
#1	.00025	.00095	.00248	.00841	-.09353	-.00318	.11978	.00265
#2	.00039	-.00048	.00263	-.00565	-.02424	.00216	.06485	.00584
#3	.00031	-.00083	.00060	.02512	.09640	-.00539	.07607	-.00060
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00355**	**.02859**	**-.00197**	**.00635**	**-.00184**	**.00648**	**-.00780**	**.00562**
Stddev	.00019	.02702	.00094	.00761	.00466	.00302	.00403	.00110
%RSD	5.3403	94.506	47.726	119.74	253.74	46.674	51.730	19.667
#1	.00349	.02662	-.00294	-.00235	-.00211	.00986	-.01231	.00668
#2	.00339	.00261	-.00192	.01175	-.00636	.00555	-.00455	.00447
#3	.00376	.05655	-.00106	.00967	.00296	.00403	-.00652	.00571
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: CCB Acquired: 6/24/2016 0:12:00 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00027	.00037	-.00205	-.00398	-.00052	.00005	-.03486
Stddev	.00199	.00049	.00712	.00235	.00010	.00013	.29949
%RSD	737.85	134.70	347.62	58.933	18.924	280.33	859.07
#1	.00244	-.00019	-.00912	-.00572	-.00063	.00006	.11572
#2	-.00149	.00054	.00512	-.00493	-.00049	-.00009	.15945
#3	-.00013	.00074	-.00215	-.00131	-.00044	.00017	-.37976
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6385.3	39685.	2013.6
Stddev	13.4	625.	15.5
%RSD	.20981	1.5744	.76988
#1	6370.9	40406.	2004.4
#2	6397.5	39350.	2031.5
#3	6387.5	39298.	2004.9

Sample Name: LLCCV Acquired: 6/24/2016 0:16:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00692	.17280	.00777	.08023	.00816	.00159	.47649	.00097
Stddev	.00020	.00275	.00190	.00358	.00068	.00009	.02580	.00026
%RSD	2.8583	1.5889	24.477	4.4629	8.3509	5.9617	5.4139	26.520
#1	.00710	.16995	.00825	.08038	.00802	.00170	.44675	.00126
#2	.00671	.17543	.00940	.08374	.00756	.00154	.48987	.00076
#3	.00694	.17303	.00568	.07658	.00890	.00153	.49285	.00090
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00419	.00395	.00595	.08433	.74272	.08170	.43560	.00898
Stddev	.00055	.00077	.00059	.03278	.15549	.00079	.02792	.00231
%RSD	13.081	19.592	9.9605	38.876	20.936	.96348	6.4091	25.784
#1	.00481	.00337	.00657	.12211	.85231	.08123	.41810	.00714
#2	.00397	.00365	.00539	.06743	.81110	.08127	.46780	.01158
#3	.00379	.00483	.00590	.06344	.56476	.08261	.42090	.00822
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00841	.43808	.01603	.80173	.00604	.08746	.00868	.86554
Stddev	.00048	.03267	.00038	.00808	.00286	.00565	.00374	.00329
%RSD	5.6889	7.4567	2.3907	1.0073	47.406	6.4564	43.147	.37989
#1	.00807	.47434	.01600	.80832	.00935	.08147	.01238	.86709
#2	.00820	.41094	.01567	.80416	.00429	.08820	.00489	.86176
#3	.00895	.42897	.01643	.79272	.00448	.09269	.00877	.86777
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: LLCCV Acquired: 6/24/2016 0:16:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41147**	**.04220**	**.01937**	**.15204**	**.00784**	**.01708**	**20.869**
Stddev	.00081	.00045	.00562	.00340	.00117	.00020	.318
%RSD	.19753	1.0715	29.021	2.2390	14.946	1.1523	1.5231
#1	.41153	.04213	.02561	.15455	.00880	.01720	21.227
#2	.41224	.04269	.01469	.15340	.00817	.01718	20.618
#3	.41062	.04179	.01783	.14816	.00653	.01685	20.762
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6476.0**	**40121.**	**2010.8**
Stddev	41.9	341.	10.1
%RSD	.64677	.85050	.50105
#1	6439.6	40107.	2010.7
#2	6521.8	39788.	2020.9
#3	6466.6	40470.	2000.7

Sample Name: LLCCV Acquired: 6/24/2016 0:19:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02136	.41629	.02112	.19510	.02080	.00407	1.0587
Stddev	.00041	.00284	.00410	.00052	.00091	.00010	.0154
%RSD	1.9202	.68293	19.397	.26774	4.3671	2.4437	1.4499
#1	.02183	.41875	.01725	.19543	.02070	.00418	1.0490
#2	.02111	.41695	.02541	.19450	.02176	.00402	1.0764
#3	.02114	.41318	.02071	.19539	.01995	.00401	1.0506
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00191	.01019	.00970	.01209	.22512	2.0312	.20560
Stddev	.00008	.00029	.00094	.00032	.02751	.1478	.00292
%RSD	3.9464	2.8921	9.6498	2.6079	12.219	7.2758	1.4184
#1	.00189	.00987	.00862	.01175	.23193	1.9599	.20462
#2	.00184	.01046	.01027	.01237	.19484	2.2011	.20331
#3	.00199	.01022	.01020	.01214	.24858	1.9325	.20889
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.0419	.02143	.01974	1.0634	.04092	1.9580	.02354
Stddev	.1407	.00617	.00054	.0338	.00119	.0082	.00145
%RSD	13.508	28.783	2.7554	3.1802	2.8968	.41842	6.1627
#1	1.1704	.01821	.01917	1.0842	.04126	1.9586	.02462
#2	1.0638	.01754	.01982	1.0815	.03960	1.9658	.02189
#3	.89149	.02854	.02024	1.0243	.04189	1.9495	.02411
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: LLCCV Acquired: 6/24/2016 0:19:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.21053**	**.03539**	**2.1191**	**1.0135**	**.10298**	**.05480**	**.37888**
Stddev	.00225	.00798	.0028	.0011	.00044	.01048	.00446
%RSD	1.0680	22.551	.13372	.10864	.43012	19.119	1.1784
#1	.21009	.02618	2.1179	1.0144	.10249	.04402	.38327
#2	.21296	.04014	2.1223	1.0138	.10312	.05543	.37434
#3	.20853	.03987	2.1170	1.0123	.10334	.06494	.37903
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.01937**	**.04203**	F **55.206**
Stddev	.00081	.00024	.600
%RSD	4.1646	.56559	1.0870
#1	.02030	.04231	55.899
#2	.01893	.04187	54.851
#3	.01888	.04193	54.868
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6500.2**	**40568.**	**2033.1**
Stddev	36.1	206.	33.6
%RSD	.55595	.50832	1.6527
#1	6459.1	40580.	1998.8
#2	6514.4	40767.	2065.9
#3	6527.1	40355.	2034.6

Element, Wavelength and Order	Date of Fit	Date of Cal.	Type of Fit	Weighting	A0	A1	A2	n (Exponent)
Ag 328.068 {103}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000236	0.051213	0.000000	1.000000
Al 308.215 {109}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000796	0.008044	0.000000	1.000000
As 189.042 {478}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000007	0.025071	0.000000	1.000000
B 249.678 {135}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000040	0.023963	0.000000	1.000000
Ba 455.403 { 74}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.044484	2.168568	0.000000	1.000000
Be 313.107 {108}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000479	0.930119	0.000000	1.000000
Ca 422.673 { 80}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.002874	0.045862	0.000000	1.000000
Cd 228.802 {447}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000038	0.570455	0.000000	1.000000
Co 228.616 {447}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000129	0.339574	0.000000	1.000000
Cr 267.716 {126}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000254	0.058743	0.000000	1.000000
Cu 224.700 {450}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000063	0.129645	0.000000	1.000000
Fe 261.187 {129}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000714	0.024119	0.000000	1.000000
K 766.490 { 44}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.006604	0.061274	0.000000	1.000000
Li 670.784 { 50}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.009415	1.520559	0.000000	1.000000
Mg 279.079 {121}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000260	0.008000	0.000000	1.000000
Mn 257.610 {131}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000470	0.289814	0.000000	1.000000
Mo 202.030 {467}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000054	0.182736	0.000000	1.000000
Na 589.592 { 57}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.047777	0.184477	0.000000	1.000000
Ni 231.604 {446}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000734	0.138990	0.000000	1.000000
P 214.914 {457}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000172	0.014319	0.000000	1.000000
Pb 220.353 {453}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.001186	0.065283	0.000000	1.000000
Sb 206.833 {463}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000368	0.031084	0.000000	1.000000
Se 196.090 {472}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000508	0.017412	0.000000	1.000000
Si 212.412 {459}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.001467	0.039573	0.000000	1.000000
Sn 189.989 {477}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000022	0.066307	0.000000	1.000000
Sr 407.771 { 83}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.001060	3.591964	0.000000	1.000000
Ti 337.280 {100}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.002630	0.122957	0.000000	1.000000
Tl 190.856 {477}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000329	0.028655	0.000000	1.000000
V 292.402 {115}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000071	0.067328	0.000000	1.000000
Y 224.306 {450}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Y 360.073 { 94}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Y 377.433 { 89}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Zn 206.200 {463}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000142	0.437417	0.000000	1.000000
Zr 339.198 { 99}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.007106	1.219765	0.000000	1.000000

Approved: June 27, 2016

Element, Wavelength and Order	Correlation	Std Error of Est	Predicted MDL	Predicted MQL	Status	Reslope		QC Norm	
						Slope	Y-int	Slope factor	Offset
Ag 328.068 {103}	0.999680	0.000003	0.002344	0.007814	OK.	1.000000	0.000000	1	0
Al 308.215 {109}	0.999950	0.000005	0.010188	0.033959	OK.	1.000000	0.000000	1	0
As 189.042 {478}	0.999909	0.000001	0.003189	0.010630	OK.	1.000000	0.000000	1	0
B 249.678 {135}	0.999961	0.000001	0.002963	0.009878	OK.	1.000000	0.000000	1	0
Ba 455.403 { 74}	0.999951	0.000136	0.001119	0.003729	OK.	1.000000	0.000000	1	0
Be 313.107 {108}	0.999766	0.000007	0.000105	0.000348	OK.	1.000000	0.000000	1	0
Ca 422.673 { 80}	0.999890	0.000043	0.043792	0.145974	OK.	1.000000	0.000000	1	0
Cd 228.802 {447}	0.999970	0.000002	0.000282	0.000940	OK.	1.000000	0.000000	1	0
Co 228.616 {447}	0.999977	0.000003	0.000454	0.001514	OK.	1.000000	0.000000	1	0
Cr 267.716 {126}	0.999978	0.000001	0.001307	0.004355	OK.	1.000000	0.000000	1	0
Cu 224.700 {450}	0.999889	0.000006	0.001405	0.004682	OK.	1.000000	0.000000	1	0
Fe 261.187 {129}	0.999330	0.000022	0.023003	0.076675	OK.	1.000000	0.000000	1	0
K 766.490 { 44}	0.999988	0.000096	0.104199	0.347330	OK.	1.000000	0.000000	1	0
Li 670.784 { 50}	0.999977	0.000100	0.004201	0.014002	OK.	1.000000	0.000000	1	0
Mg 279.079 {121}	0.999990	0.000003	0.092396	0.307987	OK.	1.000000	0.000000	1	0
Mn 257.610 {131}	0.999236	0.000036	0.002652	0.008839	OK.	1.000000	0.000000	1	0
Mo 202.030 {467}	0.999956	0.000011	0.000450	0.001500	OK.	1.000000	0.000000	1	0
Na 589.592 { 57}	0.999990	0.000264	0.032289	0.107629	OK.	1.000000	0.000000	1	0
Ni 231.604 {446}	0.999932	0.000005	0.001180	0.003935	OK.	1.000000	0.000000	1	0
P 214.914 {457}	0.999990	0.000004	0.007138	0.023792	OK.	1.000000	0.000000	1	0
Pb 220.353 {453}	0.999383	0.000007	0.003667	0.012223	OK.	1.000000	0.000000	1	0
Sb 206.833 {463}	0.999854	0.000004	0.004516	0.015054	OK.	1.000000	0.000000	1	0
Se 196.090 {472}	0.999238	0.000003	0.006704	0.022346	OK.	1.000000	0.000000	1	0
Si 212.412 {459}	0.999594	0.000036	0.002443	0.008145	OK.	1.000000	0.000000	1	0
Sn 189.989 {477}	0.999980	0.000003	0.000986	0.003287	OK.	1.000000	0.000000	1	0
Sr 407.771 { 83}	0.999952	0.000224	0.000467	0.001557	OK.	1.000000	0.000000	1	0
Ti 337.280 {100}	0.999579	0.000023	0.007453	0.024845	OK.	1.000000	0.000000	1	0
Tl 190.856 {477}	0.999831	0.000002	0.003284	0.010945	OK.	1.000000	0.000000	1	0
V 292.402 {115}	0.999969	0.000003	0.001095	0.003650	OK.	1.000000	0.000000	1	0
Y 224.306 {450}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Y 360.073 { 94}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Y 377.433 { 89}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Zn 206.200 {463}	0.999941	0.000030	0.000202	0.000675	OK.	1.000000	0.000000	1	0
Zr 339.198 { 99}	0.596446	0.010414	0.001594	0.005312	OK.	1.000000	0.000000	1	0

Approved: June 27, 2016

Sample Name: S0 Acquired: 6/25/2016 12:21:04 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00024**	**.00080**	**-.00001**	**.00004**	**.04448**	**.00048**	**.00288**
Stddev	.00008	.00004	.00007	.00002	.00170	.00005	.00169
%RSD	35.712	4.5527	977.99	54.959	3.8148	10.198	58.629
#1	-.00024	.00076	-.00006	.00003	.04635	.00042	.00096
#2	-.00015	.00081	.00007	.00006	.04405	.00050	.00353
#3	-.00032	.00083	-.00004	.00003	.04304	.00052	.00414

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00004**	**-.00013**	**.00025**	**-.00006**	**-.00071**	**-.00661**	**-.00941**
Stddev	.00006	.00005	.00003	.00023	.00026	.00355	.00965
%RSD	153.97	36.064	10.924	367.85	35.978	53.724	102.48
#1	-.00002	-.00016	.00028	-.00022	-.00093	-.00948	-.01508
#2	-.00010	-.00015	.00025	.00020	-.00043	-.00769	.00173
#3	.00001	-.00008	.00023	-.00018	-.00078	-.00264	-.01488

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00026**	**.00047**	**.00005**	**-.04778**	**-.00073**	**-.00017**	**-.00119**
Stddev	.00066	.00061	.00004	.00421	.00008	.00002	.00008
%RSD	255.00	128.77	78.942	8.8105	11.011	13.180	6.4052
#1	.00045	-.00015	.00007	-.05263	-.00065	-.00015	-.00110
#2	-.00086	.00049	.00001	-.04509	-.00082	-.00019	-.00123
#3	-.00037	.00107	.00009	-.04563	-.00073	-.00017	-.00123

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00037**	**-.00051**	**.00147**	**.00002**	**.00106**	**-.00263**	**-.00033**
Stddev	.00001	.00007	.00009	.00003	.00137	.00025	.00002
%RSD	3.7153	13.815	6.2921	126.34	130.24	9.4020	7.1687
#1	.00035	-.00043	.00157	.00002	.00203	-.00235	-.00030
#2	.00038	-.00057	.00145	-.00000	.00165	-.00283	-.00033
#3	.00037	-.00052	.00138	.00005	-.00052	-.00270	-.00035

Approved: June 27, 2016

Sample Name: S0 Acquired: 6/25/2016 12:21:04 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S
Avg	**-.00007**	**.00014**	**-.00710**
Stddev	.00004	.00005	.00324
%RSD	58.279	35.146	45.700
#1	-.00004	.00009	-.00457
#2	-.00005	.00016	-.01076
#3	-.00012	.00018	-.00597
Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9736.8**	**58329.**	**2962.8**
Stddev	24.6	275.	12.4
%RSD	.25315	.47180	.41783
#1	9719.4	58177.	2949.0
#2	9725.9	58162.	2966.1
#3	9765.0	58646.	2973.1

Approved: June 27, 2016

Sample Name: S1 Acquired: 6/25/2016 12:25:02 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	Ba4554	Be3131	Ca4226	Cd2288	Co2286
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00000**	**.00148**	**.06336**	**.00097**	**.00582**	**.00018**	**.00038**
Stddev	.00004	.00005	.00419	.00009	.00234	.00011	.00006
%RSD	838.91	3.6655	6.6080	8.8478	40.284	60.110	16.615
#1	.00001	.00145	.05987	.00105	.00569	.00026	.00043
#2	-.00004	.00145	.06800	.00097	.00822	.00006	.00031
#3	.00004	.00155	.06221	.00088	.00354	.00023	.00041

Elem	Cr2677	Cu2247	Fe2611	K_7664	Mn2576	Mo2020	Na5895
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00050**	**.00053**	**.00033**	**.01950**	**.00077**	**.00143**	**.02669**
Stddev	.00003	.00011	.00021	.00355	.00009	.00011	.00332
%RSD	5.5692	20.305	62.737	18.223	11.862	7.4791	12.449
#1	.00049	.00060	.00053	.01550	.00083	.00145	.02356
#2	.00048	.00041	.00012	.02070	.00066	.00152	.03018
#3	.00053	.00058	.00034	.02230	.00081	.00131	.02634

Elem	Ni2316	P_2149	Pb2203	Sb2068	Si2124	Sn1899	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00018**	**.00100**	**-.00086**	**.00059**	**.00297**	**.00055**	**.03302**
Stddev	.00009	.00012	.00030	.00014	.00005	.00006	.00089
%RSD	52.159	12.366	34.406	23.539	1.7045	10.953	2.6942
#1	-.00010	.00113	-.00102	.00051	.00294	.00050	.03225
#2	-.00028	.00098	-.00052	.00075	.00302	.00062	.03399
#3	-.00016	.00088	-.00105	.00050	.00293	.00055	.03280

Elem	Ti3372	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00218**	**.00050**	**.00363**	**-.00591**
Stddev	.00071	.00003	.00003	.00134
%RSD	32.781	5.1557	.77751	22.659
#1	-.00139	.00050	.00365	-.00576
#2	-.00277	.00048	.00360	-.00465
#3	-.00237	.00053	.00364	-.00732

Approved: June 27, 2016

Sample Name: S1 Acquired: 6/25/2016 12:25:02 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9839.2**	**59021.**	**2974.6**
Stddev	35.8	140.	1.9
%RSD	.36397	.23691	.06440
#1	9811.4	58880.	2976.5
#2	9826.5	59023.	2972.7
#3	9879.6	59160.	2974.8

Approved: June 27, 2016

Sample Name: S2 Acquired: 6/25/2016 12:29:00 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00011**	**.00220**	**.00018**	**.00025**	**.08112**	**.00139**	**.00931**
Stddev	.00003	.00006	.00005	.00002	.00125	.00003	.00047
%RSD	29.658	2.5782	25.201	8.0082	1.5417	1.8740	5.0396
#1	.00012	.00221	.00013	.00026	.08191	.00140	.00905
#2	.00012	.00214	.00022	.00027	.08177	.00136	.00985
#3	.00007	.00225	.00019	.00023	.07968	.00140	.00903

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00041**	**.00102**	**.00072**	**.00104**	**.00148**	**.04190**	**.01415**
Stddev	.00007	.00003	.00001	.00014	.00030	.00236	.00012
%RSD	16.350	3.0292	1.5873	13.319	20.240	5.6312	.86545
#1	.00045	.00101	.00072	.00109	.00137	.04369	.01405
#2	.00033	.00099	.00073	.00088	.00125	.03923	.01429
#3	.00045	.00105	.00071	.00115	.00182	.04279	.01411

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00109**	**.00274**	**.00298**	**.10797**	**.00048**	**.00211**	**-.00045**
Stddev	.00041	.00057	.00005	.00406	.00013	.00008	.00017
%RSD	37.607	20.931	1.7509	3.7561	26.749	3.7519	39.075
#1	.00144	.00213	.00304	.11260	.00057	.00211	-.00064
#2	.00064	.00283	.00295	.10501	.00033	.00203	-.00031
#3	.00120	.00327	.00295	.10631	.00054	.00219	-.00039

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00105**	**-.00034**	**.00448**	**.00108**	**.06082**	**-.00047**	**-.00014**
Stddev	.00006	.00004	.00009	.00004	.00131	.00089	.00005
%RSD	6.0784	12.562	2.0565	3.6593	2.1502	187.41	36.276
#1	.00109	-.00029	.00447	.00110	.05950	-.00104	-.00016
#2	.00107	-.00037	.00458	.00110	.06212	-.00093	-.00008
#3	.00097	-.00035	.00440	.00103	.06083	.00055	-.00017

Approved: June 27, 2016

Sample Name: S2 Acquired: 6/25/2016 12:29:00 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S
Avg	.00105	.00728	-.00437
Stddev	.00006	.00006	.00132
%RSD	5.7588	.85058	30.251
#1	.00102	.00735	-.00588
#2	.00112	.00727	-.00376
#3	.00101	.00722	-.00345
Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9827.0	59373.	3004.6
Stddev	11.4	114.	4.0
%RSD	.11574	.19273	.13472
#1	9814.0	59255.	3000.6
#2	9832.7	59484.	3008.7
#3	9834.5	59381.	3004.6

Approved: June 27, 2016

Sample Name: S3 Acquired: 6/25/2016 12:32:56 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.02035	.08278	.00992	.01198	2.2359	.04820	.46599	.03126
Stddev	.00007	.00029	.00007	.00002	.0069	.00016	.00236	.00006
%RSD	.36344	.34990	.73624	.13878	.30693	.33862	.50730	.19965
#1	.02029	.08301	.00987	.01198	2.2433	.04825	.46861	.03126
#2	.02044	.08286	.01000	.01199	2.2346	.04833	.46401	.03132
#3	.02033	.08245	.00988	.01196	2.2297	.04802	.46536	.03119

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.06810	.02991	.06581	.09648	3.0722	1.5250	.07962	.14725
Stddev	.00012	.00021	.00002	.00063	.0027	.0048	.00098	.00051
%RSD	.17879	.70022	.03611	.65013	.08880	.31624	1.2353	.34909
#1	.06804	.03009	.06584	.09719	3.0745	1.5268	.07965	.14783
#2	.06801	.02997	.06579	.09600	3.0692	1.5288	.07861	.14685
#3	.06824	.02968	.06581	.09626	3.0730	1.5196	.08058	.14706

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.18510	9.1987	.06966	.14157	.03222	.03693	.00649	.20808
Stddev	.00068	.0276	.00008	.00029	.00023	.00011	.00003	.00004
%RSD	.36712	.29975	.11504	.20555	.70478	.30304	.43769	.01830
#1	.18567	9.2297	.06963	.14170	.03236	.03694	.00648	.20811
#2	.18529	9.1894	.06975	.14177	.03196	.03703	.00652	.20808
#3	.18435	9.1769	.06959	.14123	.03234	.03681	.00647	.20804

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.06680	3.6291	.12115	.01313	.06767	.44396	3.5474
Stddev	.00012	.0161	.00039	.00012	.00004	.00054	.0233
%RSD	.17282	.44347	.32200	.91385	.05243	.12075	.65596
#1	.06677	3.6476	.12157	.01325	.06770	.44451	3.5656
#2	.06671	3.6220	.12107	.01301	.06768	.44393	3.5554
#3	.06693	3.6179	.12080	.01311	.06763	.44344	3.5212

Approved: June 27, 2016

Sample Name: S3 Acquired: 6/25/2016 12:32:56 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9929.1	59890.	3083.9
Stddev	18.0	90.	16.7
%RSD	.18173	.14986	.54188
#1	9944.3	59787.	3069.2
#2	9933.7	59947.	3080.5
#3	9909.1	59936.	3102.1

Approved: June 27, 2016

Sample Name: S4 Acquired: 6/25/2016 12:36:34 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.04037	.16212	.01994	.02391	4.3553	.09476	.91725	.06225
Stddev	.00010	.00009	.00010	.00008	.0282	.00010	.00714	.00008
%RSD	.26004	.05310	.51424	.32854	.64632	.10278	.77855	.12431
#1	.04040	.16220	.02005	.02399	4.3445	.09472	.91430	.06220
#2	.04046	.16203	.01991	.02383	4.3342	.09469	.91205	.06220
#3	.04026	.16211	.01985	.02390	4.3873	.09487	.92539	.06233

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.13502	.05877	.12834	.19056	6.1046	3.0186	.15914	.28946
Stddev	.00023	.00007	.00015	.00161	.0456	.0309	.00223	.00202
%RSD	.17276	.12518	.11394	.84544	.74756	1.0253	1.3995	.69732
#1	.13475	.05870	.12846	.19050	6.0869	3.0140	.15736	.28818
#2	.13515	.05878	.12838	.18897	6.0705	2.9902	.15842	.28841
#3	.13517	.05884	.12818	.19219	6.1564	3.0516	.16164	.29179

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.36324	18.368	.13725	.28094	.06365	.07367	.01336	.39092
Stddev	.00176	.146	.00020	.00054	.00019	.00012	.00006	.00048
%RSD	.48547	.79728	.14492	.19364	.30628	.15884	.42128	.12331
#1	.36486	18.303	.13728	.28153	.06355	.07380	.01339	.39147
#2	.36349	18.266	.13744	.28083	.06388	.07365	.01329	.39059
#3	.36136	18.536	.13705	.28046	.06354	.07357	.01339	.39071

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.13179	7.1439	.24230	.02560	.13357	.86773	.12255
Stddev	.00029	.0568	.00280	.00005	.00025	.00172	.00222
%RSD	.21785	.79476	1.1551	.20145	.18401	.19856	1.8114
#1	.13210	7.1181	.24103	.02555	.13357	.86947	.12020
#2	.13172	7.1045	.24036	.02563	.13332	.86769	.12285
#3	.13154	7.2090	.24551	.02564	.13382	.86603	.12461

Approved: June 27, 2016

Sample Name: S4 Acquired: 6/25/2016 12:36:34 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9795.6**	**59125.**	**3064.9**
Stddev	6.6	26.	19.7
%RSD	.06757	.04412	.64222
#1	9799.1	59120.	3071.1
#2	9799.7	59153.	3080.8
#3	9787.9	59102.	3042.9

Approved: June 27, 2016

Sample Name: ICV Acquired: 6/25/2016 12:40:09 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.40482	10.096	.40189	.50535	1.0085	.05038	10.096
Stddev	.00169	.009	.00320	.00259	.0096	.00007	.103
%RSD	.41792	.09100	.79651	.51179	.95284	.13225	1.0175
#1	.40318	10.095	.40558	.50259	.99972	.05037	10.065
#2	.40656	10.106	.40020	.50771	1.0070	.05032	10.013
#3	.40473	10.088	.39988	.50575	1.0188	.05045	10.211
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05022	.20172	.50610	.50772	4.0459	50.287	1.0069
Stddev	.00031	.00057	.00078	.00110	.0398	.526	.0059
%RSD	.61395	.28458	.15454	.21701	.98466	1.0464	.58082
#1	.05048	.20224	.50644	.50864	4.0020	49.775	1.0012
#2	.05030	.20181	.50521	.50650	4.0559	50.260	1.0067
#3	.04988	.20110	.50666	.50803	4.0798	50.827	1.0128
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	10.117	.50795	1.0200	50.068	.50832	10.053	.50460
Stddev	.152	.00312	.0050	.377	.00054	.029	.00016
%RSD	1.5029	.61369	.49437	.75311	.10596	.28628	.03077
#1	9.9425	.50440	1.0241	49.665	.50829	10.073	.50452
#2	10.219	.50922	1.0214	50.126	.50888	10.066	.50449
#3	10.191	.51024	1.0144	50.412	.50780	10.020	.50477
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: ICV Acquired: 6/25/2016 12:40:09 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2132**	**.39981**	**5.2115**	**1.0100**	**1.0094**	**1.0076**	**.50967**
Stddev	.0090	.00367	.0220	.0047	.0093	.0117	.00340
%RSD	.74499	.91900	.42193	.46646	.92212	1.1612	.66719
#1	1.2228	.39652	5.1880	1.0139	.99976	.99418	.50608
#2	1.2117	.40378	5.2316	1.0112	1.0100	1.0128	.51284
#3	1.2049	.39915	5.2149	1.0048	1.0183	1.0157	.51010
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0117**	**1.0154**	F **2.9010**
Stddev	.0015	.0015	.0057
%RSD	.14666	.14385	.19619
#1	1.0111	1.0164	2.9026
#2	1.0134	1.0160	2.8946
#3	1.0106	1.0137	2.9057
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			5.0000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9887.2**	**59604.**	**3090.1**
Stddev	2.6	87.	8.3
%RSD	.02633	.14631	.26858
#1	9888.5	59705.	3098.3
#2	9884.2	59557.	3090.3
#3	9888.9	59551.	3081.7

Approved: June 27, 2016

Sample Name: ICB Acquired: 6/25/2016 12:43:48 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00131	-.00019	.00035	.00523	.00097	.00005	-.00769	-.00024
Stddev	.00204	.00299	.00150	.00175	.00089	.00006	.00734	.00011
%RSD	155.46	1561.3	426.44	33.377	91.849	115.36	95.505	46.996
#1	.00317	-.00362	.00181	.00349	.00060	.00012	-.01556	-.00036
#2	-.00088	.00114	-.00118	.00522	.00198	.00002	-.00104	-.00017
#3	.00165	.00191	.00042	.00699	.00032	.00001	-.00646	-.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00000	.00019	.00069	.02190	.04427	.00252	.00787	.00050
Stddev	.00033	.00096	.00130	.00895	.02254	.00192	.08048	.00146
%RSD	40694.	494.95	189.51	40.889	50.920	76.061	1022.8	292.29
#1	-.00026	-.00067	-.00034	.02348	.04464	.00424	-.08504	-.00108
#2	-.00012	.00002	.00215	.01226	.06662	.00045	.05249	.00079
#3	.00038	.00123	.00025	.02996	.02154	.00288	.05615	.00179
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00437	.07087	.00070	.00285	.00118	.00678	.00870	.00121
Stddev	.00041	.01151	.00054	.00868	.00197	.00381	.00450	.00189
%RSD	9.2912	16.248	77.044	304.78	167.53	56.206	51.723	156.22
#1	.00395	.08310	.00020	-.00283	.00108	.00337	.00854	.00310
#2	.00476	.06927	.00064	-.00147	.00320	.00608	.00428	.00120
#3	.00442	.06023	.00127	.01284	-.00074	.01090	.01328	-.00068
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: ICB Acquired: 6/25/2016 12:43:48 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00063**	**.00059**	**.00116**	**.00143**	**.00111**	**.00117**	**.00345**
Stddev	.00049	.00006	.00559	.00387	.00033	.00016	.00039
%RSD	77.601	9.5579	481.47	269.75	29.521	13.916	11.209
#1	.00050	.00059	.00521	.00260	.00135	.00105	.00361
#2	.00117	.00064	.00349	-.00288	.00074	.00111	.00300
#3	.00022	.00053	-.00522	.00459	.00124	.00135	.00372
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9709.0**	**58834.**	**2997.3**
Stddev	13.6	105.	11.0
%RSD	.13963	.17854	.36570
#1	9693.7	58895.	2985.3
#2	9713.8	58895.	3006.9
#3	9719.5	58713.	2999.6

Approved: June 27, 2016

Sample Name: LLICV Acquired: 6/25/2016 12:47:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00947	.16517	.00750	.08209	.00794	.00166	.41368	.00078
Stddev	.00234	.00914	.00427	.00134	.00096	.00002	.01542	.00015
%RSD	24.742	5.5361	56.913	1.6344	12.112	1.4530	3.7271	18.963
#1	.01038	.17315	.00639	.08214	.00810	.00169	.40911	.00061
#2	.01123	.15519	.01222	.08341	.00690	.00166	.40107	.00083
#3	.00681	.16717	.00390	.08073	.00880	.00164	.43087	.00089
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00435	.00476	.00462	.09393	.88733	.08419	.44121	.00701
Stddev	.00017	.00091	.00136	.00523	.14622	.00363	.12107	.00237
%RSD	3.8803	19.165	29.360	5.5689	16.479	4.3143	27.441	33.765
#1	.00455	.00411	.00608	.08852	.75945	.08105	.39471	.00889
#2	.00425	.00438	.00436	.09433	.85580	.08335	.57864	.00779
#3	.00425	.00581	.00341	.09896	1.0467	.08816	.35028	.00435
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00866	.44156	.01715	.79893	.00705	.08442	.02038	.79310
Stddev	.00052	.01049	.00062	.00929	.00266	.00215	.00191	.00106
%RSD	5.9984	2.3750	3.6174	1.1626	37.766	2.5524	9.3696	.13338
#1	.00809	.45038	.01661	.80194	.00796	.08678	.01818	.79326
#2	.00879	.42997	.01783	.80634	.00405	.08394	.02136	.79407
#3	.00910	.44434	.01701	.78851	.00913	.08255	.02160	.79197
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LLICV Acquired: 6/25/2016 12:47:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41130**	**.04174**	**.02562**	**.15269**	**.00881**	**.01717**	**.13277**
Stddev	.00223	.00064	.00303	.00060	.00064	.00024	.00310
%RSD	.54293	1.5249	11.843	.39147	7.2679	1.3907	2.3338
#1	.41347	.04173	.02710	.15201	.00926	.01741	.13005
#2	.41142	.04111	.02213	.15292	.00808	.01715	.13211
#3	.40901	.04238	.02764	.15314	.00909	.01694	.13614
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9760.6**	**59445.**	**3032.4**
Stddev	2.7	49.	16.0
%RSD	.02790	.08308	.52659
#1	9758.1	59502.	3050.3
#2	9763.5	59418.	3027.2
#3	9760.2	59415.	3019.7

Approved: June 27, 2016

Sample Name: LLICV Acquired: 6/25/2016 12:51:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02232**	**.40482**	**.02105**	**.19795**	**.01981**	**.00414**	**.94713**	**.00209**
Stddev	.00322	.01572	.00276	.00137	.00013	.00016	.01328	.00025
%RSD	14.431	3.8838	13.088	.69255	.67950	3.8114	1.4024	12.129
#1	.02018	.38785	.01788	.19912	.01996	.00396	.94307	.00218
#2	.02076	.40770	.02252	.19644	.01971	.00427	.93634	.00229
#3	.02603	.41890	.02277	.19828	.01975	.00419	.96196	.00181
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00992**	**.00950**	**.01082**	**.20111**	**1.9275**	**.20201**	**.97307**	**.01841**
Stddev	.00043	.00028	.00021	.02025	.1498	.00564	.04542	.00186
%RSD	4.3042	2.9103	1.9731	10.069	7.7699	2.7925	4.6682	10.079
#1	.01005	.00935	.01089	.20826	1.8000	.20412	1.0187	.02052
#2	.00945	.00934	.01100	.21682	1.8900	.19562	.97274	.01702
#3	.01028	.00982	.01059	.17825	2.0924	.20630	.92781	.01771
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.01972**	**1.0340**	**.04153**	**1.9482**	**.02271**	**.20123**	**.04411**	**1.9392**
Stddev	.00048	.0096	.00031	.0042	.00204	.00295	.00361	.0022
%RSD	2.4591	.93077	.75288	.21380	8.9612	1.4657	8.1900	.11326
#1	.01969	1.0240	.04132	1.9520	.02496	.20436	.04094	1.9404
#2	.01925	1.0432	.04139	1.9437	.02217	.20085	.04804	1.9406
#3	.02022	1.0348	.04189	1.9488	.02100	.19850	.04333	1.9367
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LLICV Acquired: 6/25/2016 12:51:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.0099**	**.10045**	**.06134**	**.37714**	**.02003**	**.04134**	**.29456**
Stddev	.0057	.00069	.00459	.00090	.00046	.00030	.00329
%RSD	.56948	.68667	7.4856	.23976	2.2900	.72946	1.1169
#1	1.0145	.09991	.06229	.37809	.02052	.04128	.29317
#2	1.0117	.10021	.05634	.37629	.01997	.04166	.29219
#3	1.0035	.10123	.06538	.37705	.01961	.04107	.29832
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9725.0**	**59145.**	**3053.0**
Stddev	6.0	91.	15.9
%RSD	.06216	.15348	.52133
#1	9731.9	59116.	3042.4
#2	9721.3	59247.	3071.3
#3	9721.7	59072.	3045.2

Approved: June 27, 2016

Sample Name: ICSA Acquired: 6/25/2016 12:55:29 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00027**	**265.97**	**.00161**	**-.00823**	**.00056**	**.00011**	**243.96**	**-.00018**
Stddev	.00121	.27	.00172	.00296	.00012	.00009	1.89	.00042
%RSD	450.36	.10204	106.61	35.970	21.237	81.140	.77435	230.52
#1	-.00046	265.69	.00220	-.01033	.00047	.00017	241.81	-.00048
#2	.00103	265.97	-.00032	-.00484	.00069	.00014	245.35	.00030
#3	-.00137	266.24	.00295	-.00951	.00051	.00001	244.72	-.00037
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00099**	**-.00100**	**-.00098**	**96.848**	**.05329**	**.01035**	**248.71**	**-.00176**
Stddev	.00039	.00071	.00330	.692	.06120	.00255	1.99	.00096
%RSD	39.651	70.751	335.23	.71414	114.86	24.670	.80129	54.809
#1	-.00142	-.00119	.00263	96.056	.12150	.01230	246.41	-.00274
#2	-.00085	-.00160	-.00175	97.332	.03517	.00746	249.90	-.00082
#3	-.00068	-.00022	-.00383	97.155	.00318	.01129	249.83	-.00171
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00066**	**.04038**	**-.00089**	**.02257**	**.00214**	**-.00550**	**-.00224**	**.00879**
Stddev	.00023	.00410	.00050	.00580	.00444	.00318	.00753	.00258
%RSD	34.448	10.162	56.288	25.707	206.94	57.856	337.10	29.379
#1	-.00088	.04493	-.00121	.01613	.00721	-.00186	.00631	.00581
#2	-.00043	.03696	-.00115	.02418	.00024	-.00776	-.00511	.01037
#3	-.00066	.03925	-.00031	.02740	-.00103	-.00688	-.00791	.01019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: ICSA Acquired: 6/25/2016 12:55:29 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00053**	**.00100**	**-.00403**	**-.00373**	**-.00305**	**-.00596**	**-.01388**
Stddev	.00149	.00018	.00805	.00252	.00022	.00010	.00081
%RSD	284.28	18.415	199.81	67.478	7.0947	1.6059	5.8325
#1	.00144	.00117	.00514	-.00549	-.00300	-.00603	-.01379
#2	.00134	.00081	-.00728	-.00085	-.00329	-.00601	-.01311
#3	-.00120	.00101	-.00995	-.00485	-.00287	-.00585	-.01473
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9184.9**	**55126.**	**2960.9**
Stddev	19.5	129.	17.9
%RSD	.21199	.23455	.60311
#1	9165.3	55274.	2981.6
#2	9185.3	55069.	2950.7
#3	9204.2	55035.	2950.6

Approved: June 27, 2016

Sample Name: ICSAB Acquired: 6/25/2016 12:59:21 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51819**	**267.93**	**.25992**	**-.00749**	**.25012**	**.24499**	**247.43**	**.48653**
Stddev	.00274	.82	.00577	.00323	.00127	.00111	.96	.00069
%RSD	.52949	.30727	2.2185	43.201	.50926	.45301	.38893	.14255
#1	.51744	268.20	.26633	-.00904	.25065	.24547	247.63	.48716
#2	.52123	268.59	.25826	-.00965	.24866	.24578	246.38	.48579
#3	.51590	267.01	.25516	-.00377	.25104	.24372	248.28	.48665
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.24114**	**.25272**	**.25219**	**96.482**	**5.1931**	**.01181**	**249.25**	**.24440**
Stddev	.00029	.00219	.00111	.527	.0571	.00427	1.99	.00318
%RSD	.11831	.86807	.44033	.54597	1.0995	36.192	.80021	1.3017
#1	.24081	.25379	.25307	96.524	5.1297	.01134	249.32	.24432
#2	.24135	.25418	.25256	95.936	5.2405	.00779	247.22	.24763
#3	.24125	.25020	.25095	96.987	5.2091	.01630	251.20	.24127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00056**	**5.3241**	**.48364**	**.08508**	**.49426**	**.50665**	**.24857**	**.00554**
Stddev	.00038	.0216	.00155	.00195	.00579	.00649	.00790	.00251
%RSD	68.178	.40548	.32037	2.2888	1.1716	1.2807	3.1773	45.410
#1	-.00097	5.3151	.48290	.08463	.49996	.50736	.25719	.00434
#2	-.00022	5.3084	.48260	.08721	.48838	.49984	.24167	.00842
#3	-.00048	5.3487	.48542	.08340	.49443	.51276	.24687	.00384
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: ICSAB Acquired: 6/25/2016 12:59:21 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00018**	**.00065**	**-.00404**	**.46188**	**.25332**	**.48100**	**-.01374**
Stddev	.00129	.00010	.00503	.00318	.00107	.00194	.00218
%RSD	732.71	15.906	124.53	.68769	.42266	.40303	15.888
#1	.00146	.00072	-.00979	.45913	.25376	.48104	-.01122
#2	-.00112	.00053	-.00192	.46116	.25410	.48293	-.01498
#3	.00019	.00071	-.00042	.46536	.25210	.47905	-.01502
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9278.3**	**55833.**	**2985.5**
Stddev	15.8	186.	18.9
%RSD	.17001	.33320	.63177
#1	9263.8	55869.	2984.2
#2	9295.1	55632.	3004.9
#3	9275.9	55999.	2967.3

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 13:03:04 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39792	10.019	.39868	.49351	.99429	.04982	9.9377
Stddev	.00072	.017	.00144	.00412	.00753	.00017	.0996
%RSD	.18044	.17228	.36151	.83518	.75700	.33966	1.0022
#1	.39861	10.034	.39833	.49827	.98705	.04987	9.8876
#2	.39796	10.022	.40027	.49123	1.0021	.04997	10.052
#3	.39718	10.000	.39745	.49104	.99375	.04964	9.8731
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04945	.20002	.50114	.50617	3.9950	49.191	.99914
Stddev	.00013	.00034	.00115	.00138	.0385	.327	.00262
%RSD	.25449	.17079	.22943	.27257	.96478	.66521	.26204
#1	.04957	.20034	.49991	.50465	3.9625	48.889	.99615
#2	.04946	.20007	.50219	.50735	4.0376	49.538	1.0010
#3	.04932	.19966	.50132	.50650	3.9848	49.146	1.0003
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.9162	.49766	1.0071	49.316	.50460	9.9534	.50217
Stddev	.1345	.00161	.0027	.281	.00141	.0056	.00090
%RSD	1.3564	.32258	.27182	.56936	.27993	.05661	.17861
#1	9.8004	.49634	1.0102	49.069	.50585	9.9507	.50282
#2	10.064	.49945	1.0061	49.622	.50307	9.9496	.50115
#3	9.8846	.49718	1.0051	49.257	.50488	9.9599	.50255
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 13:03:04 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1997**	**.40476**	**5.1642**	**1.0026**	**.99432**	**.99050**	**.50900**
Stddev	.0011	.00080	.0031	.0008	.00541	.00261	.00166
%RSD	.09538	.19644	.06047	.08057	.54374	.26324	.32573
#1	1.2008	.40408	5.1631	1.0035	.99081	.99348	.51044
#2	1.1985	.40563	5.1678	1.0020	1.0005	.98942	.50718
#3	1.1997	.40457	5.1618	1.0022	.99161	.98862	.50937
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0019**	**1.0053**	F **2.8691**
Stddev	.0036	.0009	.0111
%RSD	.35558	.09150	.38634
#1	.99784	1.0055	2.8628
#2	1.0045	1.0043	2.8819
#3	1.0033	1.0060	2.8625
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9877.8**	**59691.**	**3097.9**
Stddev	6.1	104.	11.8
%RSD	.06196	.17495	.38081
#1	9880.1	59766.	3105.5
#2	9870.8	59572.	3084.3
#3	9882.4	59736.	3103.8

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 13:06:41 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00082	.00013	.00057	.00151	.00063	.00002	-.01871	-.00022
Stddev	.00151	.00746	.00362	.00174	.00068	.00003	.02739	.00016
%RSD	185.23	5748.5	636.07	114.92	108.39	116.80	146.42	72.359
#1	.00201	.00304	-.00315	-.00049	.00141	.00001	-.04795	-.00036
#2	-.00088	-.00835	.00078	.00262	.00031	.00006	.00637	-.00027
#3	.00132	.00570	.00408	.00241	.00017	.00001	-.01455	-.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00002	.00054	.00076	.03608	.09214	-.00181	-.01859	-.00071
Stddev	.00025	.00025	.00080	.01975	.18930	.00302	.07756	.00052
%RSD	1039.3	46.422	105.73	54.731	205.46	166.60	417.27	72.248
#1	.00013	.00063	.00047	.01630	-.12492	-.00468	-.09449	-.00012
#2	.00011	.00026	.00167	.05580	.17827	.00135	-.02180	-.00101
#3	-.00032	.00073	.00014	.03615	.22306	-.00212	.06053	-.00101
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00274	.04401	-.00040	.00332	.00144	-.00005	.00482	-.00649
Stddev	.00039	.03341	.00024	.00918	.00111	.00350	.00394	.00014
%RSD	14.320	75.897	59.933	276.22	77.091	6590.6	81.757	2.1187
#1	.00273	.08068	-.00054	-.00157	.00018	-.00177	.00919	-.00633
#2	.00236	.03606	-.00012	.01392	.00189	-.00236	.00153	-.00657
#3	.00314	.01530	-.00055	-.00237	.00225	.00397	.00375	-.00656
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 13:06:41 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	.00039	.00186	-.00089	-.00002	.00003	.00447
Stddev	.00029	.00029	.00050	.00195	.00109	.00004	.00129
%RSD	58.864	74.770	26.876	218.22	5660.3	163.33	28.933
#1	.00018	.00056	.00153	-.00246	.00118	.00003	.00395
#2	.00076	.00054	.00162	-.00151	-.00095	.00007	.00352
#3	.00054	.00005	.00244	.00129	-.00029	-.00002	.00594
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9714.8	59220.	3006.6
Stddev	20.8	167.	18.8
%RSD	.21426	.28151	.62581
#1	9691.7	59319.	2985.4
#2	9732.2	59028.	3021.3
#3	9720.4	59314.	3013.1

Approved: June 27, 2016

Sample Name: PBW MC Acquired: 6/25/2016 13:18:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00038**	**.00511**	**.00086**	**-.00212**	**.00115**	**.00002**	**-.06747**	**.00005**
Stddev	.00209	.01093	.00117	.00067	.00097	.00011	.03960	.00035
%RSD	552.93	213.69	135.59	31.434	84.150	461.31	58.695	692.60
#1	-.00263	-.00536	-.00048	-.00211	.00011	.00001	-.09775	.00040
#2	.00149	.00426	.00140	-.00146	.00202	-.00008	-.02265	-.00029
#3	.00000	.01644	.00166	-.00279	.00133	.00014	-.08201	.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00010**	**-.00006**	**.00212**	**.02866**	**-.06431**	**.00509**	**.02259**	**-.00086**
Stddev	.00040	.00091	.00173	.01678	.07316	.00381	.05174	.00149
%RSD	400.23	1393.8	81.515	58.547	113.75	74.903	229.03	173.59
#1	.00052	-.00104	.00296	.02246	-.13473	.00394	-.01043	-.00102
#2	.00004	.00075	.00325	.01586	-.06951	.00199	-.00402	.00071
#3	-.00026	.00010	.00013	.04766	.01130	.00934	.08222	-.00227
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00087**	**-.00687**	**.00001**	**-.00017**	**.00005**	**-.00551**	**.00627**	**.00390**
Stddev	.00030	.03350	.00031	.00833	.00114	.00506	.00234	.00374
%RSD	34.257	487.68	3374.4	4899.3	2181.8	91.834	37.291	95.869
#1	.00091	.01218	.00037	.00506	.00058	-.00133	.00544	.00713
#2	.00056	-.04555	-.00014	-.00978	.00083	-.01113	.00446	-.00020
#3	.00115	.01276	-.00019	.00420	-.00126	-.00406	.00891	.00477
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: PBW MC Acquired: 6/25/2016 13:18:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00012**	**-.00022**	**-.00094**	**-.00120**	**.00059**	**.00019**	**.00136**
Stddev	.00090	.00014	.01001	.00401	.00016	.00009	.00212
%RSD	728.27	61.668	1066.1	333.15	27.157	48.058	155.52
#1	-.00022	-.00007	.00771	-.00405	.00041	.00016	-.00108
#2	.00115	-.00034	-.01190	.00338	.00070	.00012	.00241
#3	-.00056	-.00026	.00137	-.00294	.00066	.00029	.00275
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9802.9**	**59937.**	**3028.8**
Stddev	9.7	136.	16.1
%RSD	.09849	.22735	.53011
#1	9794.7	59785.	3046.3
#2	9813.5	59979.	3014.8
#3	9800.5	60047.	3025.3

Approved: June 27, 2016

Sample Name: LCSW MC Acquired: 6/25/2016 13:22:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19140**	**4.8040**	**.19266**	**.92897**	**.49049**	**.02396**	**4.9270**	**.02412**
Stddev	.00110	.0096	.00124	.00283	.00119	.00006	.0588	.00019
%RSD	.57611	.19942	.64179	.30493	.24273	.25817	1.1941	.78757
#1	.19253	4.7978	.19184	.92575	.49050	.02398	4.9121	.02430
#2	.19133	4.7992	.19408	.93105	.49168	.02402	4.8771	.02392
#3	.19033	4.8151	.19206	.93012	.48929	.02390	4.9919	.02415
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09725**	**.24390**	**.24732**	**1.9838**	**23.909**	**.49943**	**4.9211**	**.24209**
Stddev	.00025	.00056	.00120	.0346	.055	.00048	.0513	.00093
%RSD	.26181	.22973	.48504	1.7424	.23189	.09613	1.0429	.38463
#1	.09749	.24389	.24597	1.9470	23.904	.49915	4.9459	.24237
#2	.09699	.24447	.24774	1.9888	23.968	.49916	4.8621	.24284
#3	.09727	.24335	.24825	2.0156	23.857	.49999	4.9553	.24105
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49737**	**24.381**	**.24620**	**4.6982**	**.24508**	**.58185**	**.18662**	**2.4109**
Stddev	.00101	.058	.00194	.0026	.00506	.00206	.00438	.0033
%RSD	.20378	.23960	.78649	.05522	2.0646	.35379	2.3449	.13777
#1	.49803	24.349	.24767	4.6983	.25078	.58403	.18417	2.4135
#2	.49787	24.449	.24692	4.7007	.24112	.57994	.18402	2.4071
#3	.49620	24.346	.24401	4.6956	.24334	.58158	.19167	2.4120
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LCSW MC Acquired: 6/25/2016 13:22:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48537**	**.49494**	**.49102**	**.24360**	**.48594**	**.48236**	**.01763**
Stddev	.00248	.00157	.00916	.00175	.00116	.00127	.00304
%RSD	.51037	.31733	1.8663	.71814	.23927	.26228	17.214
#1	.48605	.49357	.49562	.24211	.48702	.48137	.01600
#2	.48744	.49666	.49698	.24315	.48610	.48194	.01576
#3	.48263	.49460	.48047	.24553	.48471	.48379	.02113
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9802.2**	**59979.**	**3063.3**
Stddev	16.1	75.	6.7
%RSD	.16429	.12520	.21884
#1	9786.8	59978.	3070.3
#2	9800.7	59904.	3056.9
#3	9818.9	60054.	3062.8

Approved: June 27, 2016

Sample Name: L1606080801 Acquired: 6/25/2016 13:25:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00120	-.00080	-.00163	.02599	.00717	.00000	50.564	.00016
Stddev	.00152	.00397	.00290	.00206	.00115	.00005	.112	.00037
%RSD	125.91	493.45	178.32	7.9153	16.060	1214.4	.22137	235.03
#1	.00226	.00203	-.00429	.02577	.00588	.00002	50.480	.00003
#2	-.00053	.00090	.00147	.02405	.00757	.00004	50.521	-.00014
#3	.00188	-.00534	-.00207	.02815	.00807	-.00005	50.691	.00058
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	.00083	.00107	.26409	1.2826	.00250	14.201	.11783
Stddev	.00018	.00044	.00061	.01139	.1290	.00132	.067	.00387
%RSD	127.01	53.307	57.172	4.3115	10.056	52.928	.47527	3.2872
#1	.00035	.00131	.00089	.27352	1.4315	.00167	14.162	.12070
#2	.00008	.00073	.00057	.25144	1.2130	.00180	14.162	.11343
#3	.00000	.00044	.00175	.26731	1.2035	.00402	14.279	.11937
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00063	15.281	.00045	.01127	-.00029	-.00250	-.00036	4.8502
Stddev	.00031	.048	.00109	.00496	.00257	.00086	.00870	.0263
%RSD	48.719	.31261	245.14	44.009	886.73	34.260	2404.6	.54311
#1	.00028	15.230	.00169	.00618	-.00269	-.00223	.00526	4.8717
#2	.00080	15.325	-.00034	.01609	.00243	-.00181	.00404	4.8581
#3	.00082	15.288	-.00001	.01154	-.00061	-.00346	-.01039	4.8208
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606080801 Acquired: 6/25/2016 13:25:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00081	.16531	-.00520	-.00297	.00071	.00304	.00248
Stddev	.00021	.00042	.00147	.00169	.00080	.00019	.00016
%RSD	26.039	.25638	28.334	56.894	114.04	6.1375	6.5372
#1	.00100	.16514	-.00627	-.00112	.00132	.00299	.00242
#2	.00058	.16499	-.00352	-.00336	-.00020	.00288	.00235
#3	.00084	.16579	-.00581	-.00443	.00100	.00324	.00266
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9767.4	60311.	3068.7
Stddev	25.3	182.	4.5
%RSD	.25865	.30239	.14809
#1	9787.8	60238.	3069.0
#2	9775.3	60177.	3064.1
#3	9739.2	60519.	3073.1

Approved: June 27, 2016

Sample Name: L1606080803S Acquired: 6/25/2016 13:29:50 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19425**	**4.8313**	**.19559**	**.97188**	**.49949**	**.02391**	**55.138**	**.02433**
Stddev	.00175	.0086	.00291	.00386	.00064	.00006	.286	.00001
%RSD	.89899	.17760	1.4870	.39759	.12859	.26078	.51899	.03982
#1	.19471	4.8260	.19892	.97122	.49908	.02384	54.823	.02434
#2	.19571	4.8412	.19357	.96839	.49916	.02394	55.383	.02434
#3	.19231	4.8266	.19428	.97603	.50023	.02395	55.206	.02432
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09623**	**.24025**	**.24291**	**2.2142**	**25.706**	**.49995**	**18.850**	**.36055**
Stddev	.00052	.00032	.00081	.0203	.056	.00310	.060	.00310
%RSD	.54129	.13263	.33315	.91450	.21668	.61973	.31584	.86069
#1	.09635	.23989	.24328	2.2335	25.718	.49651	18.866	.35746
#2	.09668	.24035	.24198	2.2161	25.755	.50082	18.784	.36367
#3	.09566	.24050	.24346	2.1931	25.645	.50252	18.899	.36052
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50273**	**39.946**	**.24148**	**4.8211**	**.24377**	**.58841**	**.19283**	**7.2475**
Stddev	.00069	.089	.00226	.0078	.00185	.00329	.00168	.0253
%RSD	.13674	.22372	.93733	.16218	.76094	.55837	.87085	.34834
#1	.50346	39.846	.24394	4.8298	.24591	.59218	.19095	7.2767
#2	.50266	40.018	.24099	4.8187	.24276	.58614	.19336	7.2326
#3	.50209	39.974	.23950	4.8148	.24264	.58690	.19419	7.2333
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606080803S Acquired: 6/25/2016 13:29:50 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-07

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.48600	.66572	.49120	.24058	.48625	.48121	.00549
Stddev	.00062	.00113	.00488	.00183	.00111	.00310	.00062
%RSD	.12721	.16988	.99375	.75886	.22783	.64511	11.331
#1	.48664	.66443	.49684	.23968	.48515	.48355	.00538
#2	.48596	.66654	.48839	.24268	.48622	.48240	.00493
#3	.48540	.66619	.48838	.23938	.48737	.47769	.00616
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9640.1	59155.	3051.9
Stddev	28.9	91.	4.4
%RSD	.29949	.15374	.14285
#1	9607.9	59063.	3056.9
#2	9648.6	59158.	3049.0
#3	9663.8	59244.	3049.8

Approved: June 27, 2016

Sample Name: L1606080804SD Acquired: 6/25/2016 13:33:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.19367	4.8696	.19395	.97468	.50079	.02428	55.592	.02458
Stddev	.00103	.0189	.00351	.00126	.00119	.00004	.087	.00002
%RSD	.53221	.38806	1.8082	.12977	.23838	.15828	.15705	.06402
#1	.19313	4.8871	.19259	.97422	.49995	.02427	55.687	.02459
#2	.19486	4.8723	.19133	.97371	.50216	.02425	55.574	.02456
#3	.19302	4.8496	.19794	.97611	.50027	.02433	55.515	.02458
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09687	.24382	.24330	2.1942	25.602	.50391	18.927	.35670
Stddev	.00034	.00139	.00044	.0216	.187	.00317	.145	.00440
%RSD	.35450	.56820	.18086	.98630	.73073	.62836	.76565	1.2346
#1	.09649	.24542	.24364	2.2182	25.415	.50090	18.778	.36166
#2	.09698	.24295	.24347	2.1881	25.600	.50721	19.067	.35519
#3	.09715	.24309	.24280	2.1762	25.789	.50361	18.935	.35325
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50687	40.011	.24239	4.8324	.24609	.59184	.19199	7.3269
Stddev	.00174	.019	.00180	.0039	.00113	.00410	.00494	.0108
%RSD	.34421	.04769	.74422	.07995	.45954	.69206	2.5730	.14791
#1	.50863	40.009	.24329	4.8354	.24728	.59240	.18629	7.3276
#2	.50684	40.031	.24358	4.8338	.24503	.59563	.19457	7.3373
#3	.50514	39.993	.24032	4.8281	.24594	.58749	.19509	7.3157
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606080804SD Acquired: 6/25/2016 13:33:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-08

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.48884	.66564	.48806	.24641	.49385	.48167	.00481
Stddev	.00243	.00060	.00210	.00100	.00262	.00319	.00105
%RSD	.49731	.09007	.43005	.40687	.53028	.66320	21.781
#1	.48902	.66631	.48604	.24752	.49576	.48059	.00363
#2	.49118	.66515	.49023	.24612	.49493	.48526	.00518
#3	.48633	.66545	.48792	.24558	.49086	.47916	.00563
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9643.8	58961.	3032.4
Stddev	19.6	179.	14.4
%RSD	.20367	.30279	.47352
#1	9664.0	58941.	3022.3
#2	9624.8	58793.	3026.2
#3	9642.5	59149.	3048.8

Approved: June 27, 2016

Sample Name: L1606082203 Acquired: 6/25/2016 13:37:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00071	.03250	.00108	.03413	.76881	.00001	31.323	.00009
Stddev	.00326	.00113	.00298	.00231	.00291	.00007	.069	.00003
%RSD	461.08	3.4738	274.98	6.7818	.37861	986.76	.21967	35.456
#1	-.00056	.03174	-.00033	.03677	.76753	-.00000	31.323	.00006
#2	.00440	.03379	-.00093	.03317	.76676	.00008	31.255	.00011
#3	-.00173	.03195	.00451	.03246	.77215	-.00006	31.393	.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00298	.00799	.00175	5.1531	49.440	.10315	14.307	.17132
Stddev	.00036	.00030	.00156	.0220	.158	.00201	.082	.00563
%RSD	12.200	3.7657	89.202	.42624	.31871	1.9519	.57045	3.2873
#1	.00316	.00823	.00355	5.1330	49.302	.10173	14.214	.17528
#2	.00323	.00810	.00078	5.1498	49.405	.10227	14.338	.17381
#3	.00256	.00765	.00092	5.1766	49.612	.10546	14.368	.16487
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00045	90.168	.00369	.00788	-.00122	.00230	.00147	11.287
Stddev	.00023	.210	.00039	.00442	.00289	.00119	.00467	.027
%RSD	50.114	.23246	10.480	56.105	237.50	51.496	317.01	.23968
#1	.00031	90.009	.00403	.01267	.00108	.00337	.00397	11.318
#2	.00033	90.090	.00327	.00394	-.00447	.00102	-.00391	11.275
#3	.00071	90.405	.00377	.00704	-.00027	.00252	.00436	11.268
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082203 Acquired: 6/25/2016 13:37:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00007**	**.79119**	**.00045**	**-.00075**	**.00043**	**.00534**	**-.00006**
Stddev	.00030	.00071	.00315	.00229	.00033	.00006	.00108
%RSD	437.99	.09029	691.72	303.10	75.619	1.1956	1667.5
#1	-.00035	.79036	-.00147	.00135	.00018	.00529	-.00108
#2	-.00009	.79165	-.00125	-.00319	.00032	.00542	.00108
#3	.00024	.79155	.00408	-.00042	.00080	.00532	-.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9649.6**	**58766.**	**3005.5**
Stddev	29.5	106.	1.2
%RSD	.30621	.18076	.04053
#1	9617.1	58644.	3006.7
#2	9657.0	58831.	3005.5
#3	9674.8	58825.	3004.3

Approved: June 27, 2016

Sample Name: L1606082202 Acquired: 6/25/2016 13:40:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573973-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00163**	**.00700**	**-.00255**	**.09063**	**.14584**	**.00010**	**6.2202**	**-.00011**
Stddev	.00068	.00493	.00215	.00172	.00146	.00002	.0817	.00004
%RSD	41.927	70.458	84.324	1.8967	.99792	21.043	1.3126	39.110
#1	.00225	.01016	-.00441	.09032	.14618	.00008	6.1515	-.00007
#2	.00090	.00132	-.00302	.08908	.14425	.00013	6.1986	-.00016
#3	.00175	.00952	-.00020	.09248	.14710	.00010	6.3104	-.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00004**	**.00096**	**.00077**	**2.0756**	**1.5049**	**.02492**	**4.3992**	**.07622**
Stddev	.00041	.00002	.00085	.0056	.1629	.00329	.1257	.00201
%RSD	1133.0	2.0322	110.52	.26897	10.822	13.197	2.8567	2.6344
#1	-.00024	.00097	.00165	2.0692	1.6925	.02131	4.4123	.07724
#2	.00043	.00097	.00070	2.0781	1.3995	.02572	4.2675	.07752
#3	-.00030	.00094	-.00004	2.0795	1.4227	.02774	4.5178	.07391
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00024**	**83.246**	**.00061**	**.10079**	**.00421**	**-.00520**	**.00452**	**12.135**
Stddev	.00009	.263	.00086	.00636	.00131	.00271	.00174	.029
%RSD	36.387	.31568	140.09	6.3065	31.208	52.125	38.464	.24257
#1	-.00033	83.275	.00060	.10060	.00298	-.00301	.00254	12.154
#2	-.00015	82.970	.00147	.10724	.00559	-.00823	.00579	12.151
#3	-.00025	83.493	-.00024	.09454	.00404	-.00437	.00523	12.101
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082202 Acquired: 6/25/2016 13:40:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573973-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00029	.21054	.00087	-.00133	-.00012	.00182	.00052
Stddev	.00041	.00091	.00263	.00076	.00048	.00007	.00152
%RSD	145.47	.43126	300.38	57.105	397.63	3.7016	294.24
#1	.00007	.21002	.00237	-.00045	.00028	.00184	-.00094
#2	.00003	.21002	.00241	-.00174	.00001	.00175	.00210
#3	.00076	.21159	-.00216	-.00179	-.00065	.00188	.00040
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9729.1	59376.	3030.4
Stddev	8.6	260.	19.9
%RSD	.08835	.43738	.65516
#1	9720.3	59079.	3018.5
#2	9729.3	59490.	3053.3
#3	9737.5	59559.	3019.5

Approved: June 27, 2016

Sample Name: L1606082202PS Acquired: 6/25/2016 13:44:52 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19375**	**4.8752**	**.19613**	**1.0291**	**.62833**	**.02428**	**10.648**	**.02477**
Stddev	.00051	.0099	.00160	.0019	.00588	.00006	.170	.00011
%RSD	.26273	.20232	.81334	.18131	.93614	.24036	1.5959	.45206
#1	.19417	4.8642	.19702	1.0298	.63084	.02431	10.660	.02468
#2	.19391	4.8782	.19429	1.0305	.63254	.02431	10.812	.02489
#3	.19319	4.8832	.19709	1.0270	.62161	.02421	10.473	.02473
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09829**	**.24324**	**.24855**	**3.8486**	**25.821**	**.52335**	**8.7582**	**.31225**
Stddev	.00082	.00053	.00056	.0314	.285	.00671	.0434	.00355
%RSD	.83225	.21842	.22566	.81716	1.1038	1.2812	.49530	1.1380
#1	.09742	.24374	.24849	3.8714	25.861	.52361	8.7634	.31338
#2	.09905	.24330	.24914	3.8618	26.083	.52993	8.7125	.31510
#3	.09841	.24268	.24802	3.8128	25.517	.51653	8.7988	.30827
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50486**	**99.997**	**.24562**	**4.8721**	**.24648**	**.59017**	**.19243**	**13.471**
Stddev	.00060	.660	.00153	.0054	.00424	.00393	.00054	.013
%RSD	.11858	.66045	.62190	.11137	1.7208	.66559	.28037	.09301
#1	.50422	100.35	.24553	4.8783	.24168	.58648	.19181	13.458
#2	.50541	100.41	.24414	4.8682	.24971	.59430	.19267	13.471
#3	.50494	99.235	.24719	4.8698	.24807	.58975	.19281	13.483
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082202PS Acquired: 6/25/2016 13:44:52 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48807**	**.68216**	**.50019**	**.24323**	**.49372**	**.48957**	**.00258**
Stddev	.00205	.00533	.01089	.00117	.00097	.00018	.00106
%RSD	.41999	.78112	2.1777	.48173	.19621	.03662	41.003
#1	.49029	.68537	.51266	.24188	.49454	.48951	.00281
#2	.48770	.68509	.49539	.24396	.49265	.48978	.00349
#3	.48624	.67601	.49252	.24385	.49397	.48943	.00142
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9652.9**	**58738.**	**3022.5**
Stddev	16.4	44.	32.4
%RSD	.17031	.07498	1.0716
#1	9668.7	58757.	3005.1
#2	9654.1	58688.	3002.5
#3	9635.9	58769.	3059.8

Approved: June 27, 2016

Sample Name: L1606082202SDL Acquired: 6/25/2016 13:48:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00041	-.00874	.00232	.01966	.02846	.00012	1.1981	-.00020
Stddev	.00102	.00520	.00129	.00166	.00164	.00003	.0364	.00029
%RSD	249.96	59.552	55.713	8.4558	5.7787	28.080	3.0406	145.60
#1	-.00049	-.00387	.00244	.02114	.03034	.00013	1.2234	-.00054
#2	.00065	-.00812	.00098	.01996	.02778	.00008	1.1564	-.00008
#3	-.00139	-.01423	.00356	.01786	.02727	.00014	1.2146	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00018	.00097	.00049	.42314	.31666	.00414	.84994	.01478
Stddev	.00036	.00092	.00031	.02269	.03395	.00078	.10125	.00159
%RSD	202.36	94.832	64.039	5.3615	10.720	18.849	11.912	10.770
#1	-.00059	-.00007	.00031	.40193	.27970	.00483	.89447	.01388
#2	-.00002	.00128	.00031	.42043	.32384	.00429	.92129	.01385
#3	.00008	.00169	.00086	.44706	.34645	.00330	.73406	.01662
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00015	16.797	-.00010	.02573	.00379	-.00654	-.00062	2.4606
Stddev	.00016	.116	.00075	.00347	.00211	.00206	.00121	.0014
%RSD	105.18	.68795	766.63	13.470	55.760	31.494	195.90	.05654
#1	-.00025	16.870	-.00048	.02714	.00159	-.00786	-.00167	2.4590
#2	-.00024	16.856	-.00058	.02827	.00580	-.00417	-.00090	2.4616
#3	.00003	16.663	.00077	.02178	.00399	-.00761	.00071	2.4611
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082202SDL Acquired: 6/25/2016 13:48:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00034**	**.04233**	**.00225**	**-.00298**	**.00005**	**.00069**	**.00014**
Stddev	.00035	.00076	.00380	.00312	.00068	.00008	.00023
%RSD	102.54	1.7878	169.10	104.78	1298.3	11.967	160.88
#1	-.00033	.04309	.00306	.00056	-.00034	.00074	-.00011
#2	.00000	.04232	.00558	-.00534	-.00035	.00059	.00032
#3	-.00071	.04157	-.00189	-.00415	.00084	.00073	.00022
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10112.**	**61959.**	**3102.5**
Stddev	25.	335.	21.2
%RSD	.24804	.54051	.68302
#1	10140.	62322.	3081.9
#2	10091.	61890.	3101.4
#3	10107.	61663.	3124.2

Approved: June 27, 2016

Sample Name: L1606082202SDL Acquired: 6/25/2016 13:52:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00014**	**-.00201**	**.00162**	**.00506**	**.00674**	**.00009**	**.22194**	**.00006**
Stddev	.00098	.00612	.00099	.00223	.00036	.00007	.03121	.00010
%RSD	701.06	303.84	61.332	44.136	5.3564	71.933	14.062	159.17
#1	-.00091	-.00825	.00149	.00575	.00715	.00008	.25227	.00003
#2	-.00048	-.00177	.00268	.00256	.00663	.00004	.22363	-.00001
#3	.00097	.00398	.00070	.00686	.00645	.00017	.18992	.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00007**	**.00003**	**.00093**	**.10294**	**.07211**	**-.00053**	**.11677**	**.00303**
Stddev	.00044	.00176	.00026	.00844	.06933	.00190	.04489	.00266
%RSD	660.12	5191.6	27.598	8.2017	96.144	360.65	38.441	87.562
#1	-.00052	.00076	.00092	.11264	.12436	-.00025	.12873	.00344
#2	.00037	-.00197	.00068	.09726	-.00654	-.00255	.06712	.00020
#3	-.00005	.00131	.00120	.09891	.09851	.00122	.15447	.00546
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00046**	**3.4626**	**-.00095**	**.01102**	**.00309**	**-.00448**	**.00254**	**.52285**
Stddev	.00009	.0579	.00053	.00740	.00233	.00663	.00646	.00911
%RSD	18.881	1.6714	55.686	67.190	75.440	147.88	254.92	1.7427
#1	-.00041	3.4180	-.00056	.01783	.00042	-.00172	-.00323	.51493
#2	-.00042	3.4418	-.00074	.01207	.00473	-.01205	.00952	.52082
#3	-.00056	3.5280	-.00155	.00314	.00412	.00032	.00131	.53281
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082202SDL Acquired: 6/25/2016 13:52:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00039**	**.00876**	**-.00160**	**-.00170**	**.00064**	**.00048**	**.00025**
Stddev	.00024	.00020	.00243	.00072	.00151	.00001	.00052
%RSD	61.212	2.3382	151.62	42.632	234.58	1.6034	205.72
#1	-.00036	.00876	-.00015	-.00219	-.00097	.00048	.00032
#2	-.00064	.00897	-.00025	-.00087	.00088	.00049	-.00030
#3	-.00017	.00856	-.00441	-.00204	.00202	.00048	.00073
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10010.**	**61091.**	**3050.7**
Stddev	14.	94.	16.7
%RSD	.14368	.15455	.54708
#1	9999.9	61099.	3067.0
#2	10027.	60993.	3033.7
#3	10004.	61182.	3051.5

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 13:56:22 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39703	9.8823	.39208	.48360	.99649	.04877	9.9442
Stddev	.00097	.0185	.00108	.00155	.00485	.00012	.0391
%RSD	.24464	.18766	.27635	.32015	.48698	.23806	.39293
#1	.39651	9.9033	.39094	.48432	.99107	.04885	9.8991
#2	.39815	9.8682	.39222	.48182	1.0004	.04883	9.9670
#3	.39642	9.8754	.39309	.48466	.99796	.04864	9.9665
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04952	.19590	.48861	.49872	3.9553	49.086	.99832
Stddev	.00026	.00028	.00039	.00131	.0110	.181	.00236
%RSD	.52554	.14170	.07934	.26171	.27922	.36920	.23624
#1	.04970	.19602	.48888	.50017	3.9554	48.878	.99571
#2	.04922	.19610	.48816	.49764	3.9662	49.210	1.0003
#3	.04964	.19559	.48877	.49836	3.9441	49.170	.99895
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.6989	.49966	.99513	49.378	.49577	9.7713	.49937
Stddev	.0286	.00128	.00487	.189	.00112	.0077	.00125
%RSD	.29470	.25687	.48908	.38358	.22653	.07875	.25024
#1	9.6936	.49933	.99926	49.226	.49626	9.7672	.49899
#2	9.6733	.50107	.99636	49.590	.49656	9.7802	.50076
#3	9.7297	.49857	.98976	49.317	.49448	9.7665	.49835
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 13:56:22 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1744**	**.39516**	**5.0749**	**.97934**	**.99997**	**.99747**	**.49995**
Stddev	.0050	.00337	.0045	.00061	.00261	.00842	.00315
%RSD	.42778	.85272	.08873	.06267	.26098	.84451	.63052
#1	1.1693	.39868	5.0732	.98005	.99777	.99280	.50104
#2	1.1793	.39197	5.0800	.97900	1.0028	1.0072	.49639
#3	1.1748	.39484	5.0715	.97897	.99928	.99242	.50240
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.98727**	**.98962**	F **2.8790**
Stddev	.00224	.00122	.0125
%RSD	.22699	.12332	.43408
#1	.98696	.98924	2.8649
#2	.98519	.99098	2.8888
#3	.98965	.98864	2.8834
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9881.8**	**60025.**	**3051.6**
Stddev	7.5	117.	13.0
%RSD	.07613	.19437	.42745
#1	9874.3	60038.	3066.6
#2	9889.3	60134.	3043.0
#3	9881.9	59902.	3045.3

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 14:00:01 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00037	.00384	.00171	.00288	.00164	.00016	-.02160	-.00002
Stddev	.00056	.00607	.00179	.00157	.00101	.00008	.03268	.00008
%RSD	153.00	158.06	104.81	54.496	61.686	49.488	151.32	357.31
#1	-.00082	.00133	.00363	.00341	.00099	.00008	-.05876	.00002
#2	-.00054	.01077	.00142	.00412	.00280	.00016	-.00868	-.00011
#3	.00026	-.00057	.00008	.00112	.00111	.00023	.00265	.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	.00089	.00145	.02539	.05830	-.00178	-.00516	-.00289
Stddev	.00025	.00059	.00085	.02878	.16368	.00331	.04118	.00267
%RSD	184.36	65.782	58.421	113.33	280.73	185.62	797.74	92.366
#1	-.00007	.00150	.00078	.04856	-.06991	-.00470	-.03370	-.00034
#2	.00042	.00086	.00240	.03444	.24267	-.00246	.04205	-.00265
#3	.00006	.00033	.00117	-.00682	.00216	.00181	-.02383	-.00566
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00312	.01750	-.00036	.00463	.00110	.00408	.00171	-.00349
Stddev	.00035	.01788	.00074	.00054	.00119	.00430	.00409	.00216
%RSD	11.208	102.15	207.33	11.581	107.78	105.36	238.51	61.802
#1	.00276	.00856	.00044	.00497	-.00027	-.00057	.00622	-.00129
#2	.00346	.03809	-.00047	.00401	.00175	.00791	-.00177	-.00560
#3	.00314	.00586	-.00104	.00491	.00181	.00489	.00070	-.00357
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 14:00:01 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00015	.00055	-.00034	-.00341	.00088	.00013	.00321
Stddev	.00088	.00013	.00661	.00148	.00029	.00021	.00123
%RSD	577.33	24.369	1964.7	43.438	32.521	159.30	38.279
#1	-.00076	.00044	.00453	-.00437	.00106	-.00009	.00194
#2	.00100	.00070	.00232	-.00416	.00055	.00032	.00329
#3	.00022	.00052	-.00786	-.00171	.00104	.00016	.00439
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9674.1	59179.	2989.6
Stddev	10.8	80.	18.0
%RSD	.11204	.13474	.60268
#1	9680.4	59116.	2972.7
#2	9661.6	59269.	3008.6
#3	9680.2	59153.	2987.4

Approved: June 27, 2016

Sample Name: L1606082302 Acquired: 6/25/2016 14:03:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00010**	**.01235**	**.00398**	**.02065**	**7.6114**	**.00014**	**242.99**
Stddev	.00016	.00738	.00092	.00248	.0086	.00009	.50
%RSD	156.01	59.723	23.024	12.009	.11234	64.318	.20416
#1	-.00023	.01740	.00411	.01789	7.6202	.00016	242.85
#2	-.00014	.01576	.00483	.02137	7.6109	.00023	243.54
#3	.00007	.00389	.00301	.02270	7.6032	.00004	242.58
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00045**	**.05380**	**.00122**	**.00261**	**4.4999**	**3.4068**	**.13022**
Stddev	.00043	.00087	.00035	.00011	.0119	.0234	.00431
%RSD	95.240	1.6084	28.402	4.2649	.26410	.68699	3.3105
#1	-.00003	.05465	.00142	.00253	4.4869	3.3837	.13493
#2	.00060	.05384	.00143	.00274	4.5026	3.4062	.12925
#3	.00078	.05292	.00082	.00256	4.5101	3.4305	.12648
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**186.55**	**.80417**	**.00043**	F **676.67**	**.01798**	**.04840**	**.00356**
Stddev	.43	.00344	.00029	5.78	.00081	.00767	.00232
%RSD	.22901	.42830	67.402	.85481	4.4962	15.839	65.024
#1	186.94	.80321	.00050	676.18	.01876	.05723	.00107
#2	186.09	.80800	.00067	682.68	.01714	.04349	.00566
#3	186.63	.80131	.00011	671.15	.01803	.04447	.00396
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 27, 2016

Sample Name: L1606082302 Acquired: 6/25/2016 14:03:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00329	-.00499	17.541	.00002	5.8024	F -.03279	-.00373
Stddev	.00417	.00266	.115	.00016	.0077	.00249	.00217
%RSD	126.77	53.331	.65409	771.83	.13255	7.6037	58.233
#1	-.00765	-.00207	17.622	-.00016	5.8086	-.03219	-.00297
#2	-.00289	-.00563	17.591	.00008	5.8048	-.03065	-.00619
#3	.00066	-.00727	17.409	.00014	5.7938	-.03553	-.00205
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	-.00105	.00822	.00588
Stddev	.00081	.00012	.00100
%RSD	77.437	1.5051	16.931
#1	-.00020	.00826	.00695
#2	-.00112	.00832	.00498
#3	-.00182	.00808	.00571
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9001.2	54623.	2943.0
Stddev	12.7	293.	11.3
%RSD	.14121	.53595	.38250
#1	8986.8	54359.	2937.9
#2	9010.8	54574.	2935.3
#3	9006.1	54938.	2955.9

Approved: June 27, 2016

Sample Name: L1606082303 Acquired: 6/25/2016 14:07:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00097**	**.07689**	**.00082**	**.01317**	**4.0611**	**.00008**	**194.35**
Stddev	.00087	.00794	.00214	.00140	.0013	.00002	.19
%RSD	89.857	10.324	259.83	10.655	.03277	29.564	.09947
#1	.00177	.08422	.00140	.01160	4.0625	.00007	194.28
#2	.00005	.07798	.00261	.01363	4.0599	.00006	194.19
#3	.00109	.06846	-.00155	.01429	4.0610	.00010	194.56
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00041**	**.00720**	**.00128**	**.00172**	**.48254**	**1.5553**	**.05921**
Stddev	.00016	.00037	.00008	.00037	.03011	.0518	.00214
%RSD	37.982	5.1555	6.4501	21.255	6.2390	3.3303	3.6194
#1	.00025	.00709	.00125	.00159	.48289	1.5549	.05878
#2	.00056	.00690	.00137	.00144	.45226	1.5037	.05732
#3	.00044	.00762	.00121	.00214	.51247	1.6073	.06154
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**110.52**	**2.3053**	**-.00008**	F **614.40**	**.00312**	**.00621**	**.00333**
Stddev	.38	.0070	.00038	4.80	.00050	.00738	.00101
%RSD	.34277	.30352	457.85	.78181	15.943	118.75	30.264
#1	110.96	2.3090	-.00052	609.47	.00307	-.00184	.00218
#2	110.30	2.2972	.00010	614.65	.00363	.01264	.00406
#3	110.30	2.3096	.00018	619.07	.00264	.00783	.00375
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 27, 2016

Sample Name: L1606082303 Acquired: 6/25/2016 14:07:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00131	-.00398	12.572	-.00061	3.7282	-.02186	-.00195
Stddev	.00101	.00138	.069	.00058	.0049	.00737	.00114
%RSD	77.475	34.690	.54565	95.484	.12996	33.713	58.500
#1	-.00178	-.00242	12.599	.00005	3.7247	-.01374	-.00282
#2	-.00014	-.00503	12.623	-.00084	3.7262	-.02372	-.00066
#3	-.00200	-.00451	12.494	-.00105	3.7337	-.02812	-.00237
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	-.00017	.00519	.00587
Stddev	.00034	.00017	.00146
%RSD	196.61	3.3392	24.813
#1	.00020	.00538	.00454
#2	-.00026	.00515	.00565
#3	-.00046	.00504	.00743
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9113.2	55170.	2962.5
Stddev	17.2	284.	10.8
%RSD	.18895	.51491	.36572
#1	9120.9	54953.	2952.0
#2	9093.5	55065.	2973.6
#3	9125.3	55492.	2961.9

Approved: June 27, 2016

Sample Name: L1606082304 Acquired: 6/25/2016 14:12:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00108**	**.00282**	**.00375**	**.01195**	**1.2873**	**.00004**	**44.812**	**-.00032**
Stddev	.00081	.00175	.00233	.00191	.0068	.00003	.140	.00033
%RSD	75.442	62.066	62.147	15.958	.52861	68.482	.31190	102.13
#1	-.00032	.00374	.00333	.01352	1.2889	.00004	44.884	-.00060
#2	-.00194	.00392	.00166	.00983	1.2799	.00001	44.651	.00004
#3	-.00098	.00080	.00626	.01250	1.2932	.00007	44.900	-.00039
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00062**	**-.00235**	**.00403**	**80.696**	**3.0351**	**.04425**	**27.190**	**1.4212**
Stddev	.00024	.00090	.00186	.288	.0674	.00194	.251	.0124
%RSD	38.213	38.227	46.192	.35704	2.2197	4.3802	.92249	.86964
#1	.00059	-.00132	.00534	80.883	2.9579	.04406	26.967	1.4224
#2	.00088	-.00276	.00485	80.364	3.0650	.04628	27.141	1.4083
#3	.00040	-.00297	.00190	80.840	3.0823	.04242	27.462	1.4329
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00083**	**136.40**	**.00117**	**.05473**	**.00702**	**-.00148**	**-.00406**	**16.093**
Stddev	.00029	.62	.00056	.00919	.00288	.00326	.00473	.052
%RSD	35.125	.45656	47.701	16.792	41.072	220.80	116.61	.32014
#1	-.00064	136.42	.00054	.06475	.00887	-.00465	-.00925	16.145
#2	-.00116	135.77	.00135	.05277	.00848	.00187	.00003	16.091
#3	-.00068	137.01	.00161	.04668	.00370	-.00165	-.00296	16.042
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082304 Acquired: 6/25/2016 14:12:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00141**	**1.0265**	**.02088**	**-.00129**	**-.00025**	**.01157**	**-.01042**
Stddev	.00052	.0054	.00601	.00283	.00173	.00020	.00036
%RSD	36.987	.52659	28.762	219.55	704.92	1.7611	3.4676
#1	.00201	1.0281	.02616	-.00374	-.00225	.01176	-.01080
#2	.00117	1.0205	.01435	.00181	.00074	.01159	-.01039
#3	.00105	1.0309	.02213	-.00193	.00077	.01136	-.01008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9584.5**	**58328.**	**2992.4**
Stddev	11.8	87.	5.0
%RSD	.12317	.14881	.16820
#1	9573.1	58228.	2986.6
#2	9583.6	58387.	2995.2
#3	9596.7	58368.	2995.4

Approved: June 27, 2016

Sample Name: L1606114801 Acquired: 6/25/2016 14:15:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00263	.73135	.00385	.12210	.08262	.00020	27.646	.00018
Stddev	.00073	.00428	.00128	.00118	.00156	.00004	.096	.00015
%RSD	27.764	.58467	33.324	.96789	1.8866	19.880	.34678	80.450
#1	.00220	.73627	.00263	.12346	.08403	.00016	27.689	.00009
#2	.00223	.72927	.00372	.12147	.08287	.00024	27.713	.00035
#3	.00348	.72851	.00519	.12136	.08095	.00020	27.536	.00011
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00054	.00138	.00319	.50952	4.5430	.00820	13.834	.02681
Stddev	.00058	.00089	.00080	.00314	.0592	.00406	.158	.00190
%RSD	107.78	64.361	25.113	.61565	1.3042	49.473	1.1398	7.0691
#1	-.00010	.00066	.00412	.50590	4.4785	.00685	13.879	.02807
#2	.00067	.00110	.00265	.51133	4.5951	.01276	13.964	.02463
#3	.00103	.00237	.00282	.51133	4.5554	.00499	13.658	.02774
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00124	65.018	.00157	.05465	.00287	-.00148	.00632	4.3760
Stddev	.00020	.232	.00022	.00221	.00233	.00288	.00218	.0207
%RSD	16.215	.35629	13.957	4.0445	81.267	194.34	34.507	.47244
#1	.00141	65.009	.00133	.05367	.00373	-.00288	.00878	4.3946
#2	.00130	65.254	.00176	.05719	.00465	-.00339	.00555	4.3795
#3	.00102	64.791	.00160	.05310	.00023	.00183	.00463	4.3537
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606114801 Acquired: 6/25/2016 14:15:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00055**	**.26681**	**.01202**	**-.00269**	**.00480**	**.00516**	**.00567**
Stddev	.00125	.00125	.00081	.00070	.00038	.00011	.00071
%RSD	227.21	.46986	6.7372	25.927	7.9446	2.0641	12.451
#1	-.00072	.26656	.01275	-.00259	.00452	.00521	.00488
#2	.00178	.26818	.01115	-.00205	.00524	.00504	.00622
#3	.00058	.26571	.01216	-.00344	.00465	.00524	.00592
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9685.5**	**59237.**	**3013.9**
Stddev	15.3	70.	21.8
%RSD	.15804	.11840	.72168
#1	9667.9	59212.	3004.6
#2	9692.8	59182.	2998.3
#3	9695.8	59316.	3038.7

Approved: June 27, 2016

Sample Name: L1606114802 Acquired: 6/25/2016 14:19:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00086	.02576	.00392	.13054	.14235	.00004	53.032	.00013
Stddev	.00161	.00264	.00128	.00304	.00093	.00006	.088	.00011
%RSD	186.85	10.249	32.777	2.3320	.65273	165.85	.16643	81.772
#1	-.00048	.02479	.00505	.13055	.14272	-.00002	52.942	.00017
#2	.00053	.02374	.00418	.13357	.14130	.00011	53.037	.00022
#3	-.00263	.02875	.00252	.12748	.14305	.00003	53.119	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00081	.00416	.00255	.05069	2.7172	.00979	45.232	.00619
Stddev	.00057	.00090	.00135	.00895	.0969	.00331	.213	.00128
%RSD	70.201	21.666	53.148	17.648	3.5679	33.803	.47142	20.757
#1	.00109	.00488	.00157	.05068	2.6091	.01132	44.992	.00715
#2	.00016	.00315	.00198	.05964	2.7964	.01206	45.307	.00668
#3	.00119	.00444	.00409	.04175	2.7463	.00599	45.398	.00473
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	113.30	.00198	.03363	.00397	-.00032	.00108	5.7426
Stddev	.00005	.15	.00097	.00220	.00496	.00242	.00246	.0270
%RSD	9.2544	.13423	48.767	6.5321	124.89	765.53	227.39	.46957
#1	.00062	113.13	.00098	.03204	.00115	-.00095	-.00128	5.7131
#2	.00062	113.39	.00206	.03614	.00107	.00236	.00363	5.7660
#3	.00052	113.39	.00290	.03272	.00971	-.00236	.00090	5.7486
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606114802 Acquired: 6/25/2016 14:19:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00015**	**.36906**	**-.00370**	**-.00072**	**.00779**	**.00692**	**.00331**
Stddev	.00028	.00064	.00411	.00265	.00075	.00021	.00138
%RSD	191.03	.17400	110.98	366.07	9.6620	3.0569	41.716
#1	-.00017	.36882	.00101	-.00049	.00853	.00671	.00489
#2	.00034	.36857	-.00557	.00180	.00703	.00713	.00270
#3	.00027	.36978	-.00654	-.00348	.00781	.00694	.00235
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9627.8**	**58951.**	**3062.4**
Stddev	26.1	26.	1.9
%RSD	.27151	.04447	.06237
#1	9657.6	58927.	3060.7
#2	9617.0	58947.	3062.0
#3	9608.7	58979.	3064.5

Approved: June 27, 2016

Sample Name: L1606115001 Acquired: 6/25/2016 14:23:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00134	.94860	.00623	.13672	.10593	.00014	35.066	.00002
Stddev	.00163	.01243	.00298	.00233	.00120	.00009	.208	.00013
%RSD	122.05	1.3105	47.819	1.7071	1.1370	65.235	.59262	741.62
#1	-.00020	.96246	.00332	.13853	.10456	.00018	34.826	.00011
#2	.00116	.93843	.00928	.13754	.10682	.00004	35.168	-.00013
#3	.00306	.94490	.00609	.13408	.10640	.00021	35.202	.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00095	.00115	.00402	.75423	2.6231	.00517	21.361	.05361
Stddev	.00038	.00077	.00076	.00983	.0329	.00134	.091	.00159
%RSD	39.586	67.102	19.015	1.3028	1.2541	25.927	.42723	2.9705
#1	.00055	.00114	.00333	.74497	2.6490	.00636	21.310	.05382
#2	.00102	.00192	.00389	.75319	2.6342	.00372	21.307	.05192
#3	.00129	.00038	.00484	.76454	2.5861	.00543	21.467	.05508
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00320	114.16	.00202	.11444	.00566	-.00032	.00352	5.0408
Stddev	.00062	.45	.00043	.00615	.00453	.00176	.00035	.0084
%RSD	19.542	.39163	21.028	5.3725	80.076	549.83	10.037	.16741
#1	.00290	113.69	.00211	.12072	.00935	.00043	.00314	5.0497
#2	.00278	114.20	.00240	.11417	.00060	.00094	.00356	5.0330
#3	.00392	114.58	.00156	.10843	.00702	-.00233	.00385	5.0397
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115001 Acquired: 6/25/2016 14:23:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00003**	**.24066**	**.01417**	**-.00245**	**.00766**	**.00892**	**.00404**
Stddev	.00021	.00154	.00348	.00225	.00065	.00011	.00127
%RSD	805.05	.63820	24.568	91.853	8.5297	1.1926	31.485
#1	.00024	.23889	.01059	-.00116	.00840	.00901	.00275
#2	.00001	.24171	.01754	-.00114	.00737	.00895	.00530
#3	-.00018	.24137	.01438	-.00504	.00719	.00880	.00407
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9688.5**	**58932.**	**3057.0**
Stddev	25.8	93.	13.7
%RSD	.26652	.15746	.44955
#1	9698.0	58835.	3047.2
#2	9708.3	59020.	3072.7
#3	9659.3	58940.	3051.0

Approved: June 27, 2016

Sample Name: L1606115002 Acquired: 6/25/2016 14:27:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00203	.48878	.00167	.06520	.04585	.00006	17.152	.00001
Stddev	.00268	.01554	.00141	.00090	.00204	.00009	.173	.00009
%RSD	131.70	3.1791	84.852	1.3869	4.4598	167.56	1.0064	934.85
#1	.00187	.50660	.00286	.06544	.04801	.00015	17.256	-.00009
#2	-.00056	.47801	.00204	.06596	.04394	-.00004	16.953	.00007
#3	.00479	.48175	.00010	.06420	.04559	.00005	17.247	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00031	.00168	.00243	.39846	4.7446	.00319	5.4375	.01729
Stddev	.00031	.00066	.00184	.00667	.1185	.00289	.0385	.00153
%RSD	100.89	39.152	75.822	1.6749	2.4966	90.823	.70870	8.8325
#1	.00065	.00096	.00122	.39954	4.7392	.00001	5.4553	.01884
#2	.00006	.00183	.00151	.39131	4.6289	.00567	5.3932	.01727
#3	.00020	.00224	.00454	.40452	4.8657	.00387	5.4638	.01578
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00028	38.880	.00010	.07228	.00433	-.00020	.00027	1.5503
Stddev	.00033	.402	.00056	.00820	.00375	.00324	.00590	.0251
%RSD	119.49	1.0347	568.40	11.341	86.648	1601.4	2189.5	1.6195
#1	-.00010	39.210	-.00001	.06867	.00198	.00127	.00262	1.5658
#2	.00053	38.432	.00070	.06650	.00235	-.00392	-.00644	1.5637
#3	.00040	38.999	-.00040	.08166	.00865	.00205	.00464	1.5213
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115002 Acquired: 6/25/2016 14:27:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00004	.09712	.00701	-.00122	.00599	.00508	.00239
Stddev	.00062	.00057	.00249	.00352	.00080	.00018	.00169
%RSD	1544.5	.58817	35.547	288.53	13.353	3.5369	70.760
#1	.00058	.09776	.00445	-.00524	.00624	.00503	.00322
#2	-.00064	.09666	.00715	.00026	.00510	.00528	.00044
#3	.00018	.09693	.00942	.00132	.00664	.00493	.00349
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9845.3	59884.	3057.5
Stddev	61.7	141.	28.2
%RSD	.62651	.23463	.92251
#1	9792.5	59732.	3034.3
#2	9830.2	60008.	3088.9
#3	9913.1	59914.	3049.2

Approved: June 27, 2016

Sample Name: L1606115102 Acquired: 6/25/2016 14:31:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00159	.14924	.00211	.10945	.07344	.00012	35.265	.00011
Stddev	.00084	.00304	.00150	.00214	.00141	.00005	.199	.00019
%RSD	53.011	2.0340	71.147	1.9507	1.9218	40.963	.56552	183.34
#1	.00069	.14626	.00205	.10856	.07211	.00007	35.174	.00027
#2	.00171	.15233	.00364	.10790	.07329	.00017	35.493	-.00011
#3	.00236	.14912	.00064	.11188	.07492	.00013	35.127	.00016
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00052	.00053	.00363	.24070	2.2004	.00754	18.847	.05800
Stddev	.00037	.00064	.00076	.00439	.0134	.00192	.227	.00368
%RSD	71.691	119.20	20.837	1.8247	.61086	25.451	1.2033	6.3535
#1	.00009	-.00001	.00323	.23564	2.1949	.00870	18.615	.05962
#2	.00071	.00038	.00450	.24286	2.2158	.00532	19.068	.05378
#3	.00075	.00124	.00316	.24359	2.1907	.00860	18.858	.06059
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00267	93.416	.00066	.05411	.00356	-.00299	.00235	4.9794
Stddev	.00025	.232	.00043	.00846	.00271	.00266	.00452	.0235
%RSD	9.2843	.24821	65.257	15.640	76.210	88.939	192.50	.47215
#1	.00239	93.189	.00082	.06384	.00043	-.00008	-.00245	4.9968
#2	.00274	93.652	.00099	.04849	.00522	-.00529	.00653	4.9888
#3	.00288	93.408	.00017	.05000	.00501	-.00360	.00296	4.9527
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115102 Acquired: 6/25/2016 14:31:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00015	.22720	-.00049	-.00434	.00477	.01199	.00282
Stddev	.00094	.00191	.00341	.00316	.00052	.00018	.00083
%RSD	624.90	.83946	689.70	72.807	10.840	1.4772	29.345
#1	.00091	.22526	.00103	-.00351	.00496	.01201	.00271
#2	-.00090	.22907	-.00440	-.00783	.00516	.01216	.00370
#3	.00043	.22727	.00189	-.00168	.00418	.01181	.00206
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9596.9	58682.	3041.4
Stddev	35.1	192.	10.5
%RSD	.36599	.32727	.34598
#1	9558.5	58834.	3032.3
#2	9605.0	58747.	3039.1
#3	9627.4	58466.	3052.9

Approved: June 27, 2016

Sample Name: L1606115105 Acquired: 6/25/2016 14:35:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00046	.08611	.00385	.13923	.09101	.00009	44.089	.00002
Stddev	.00149	.00500	.00146	.00310	.00028	.00009	.047	.00004
%RSD	320.61	5.8036	37.928	2.2254	.30236	102.05	.10587	146.93
#1	-.00016	.08678	.00262	.14146	.09069	.00015	44.139	.00006
#2	-.00061	.09074	.00347	.13569	.09114	.00014	44.082	.00002
#3	.00216	.08081	.00547	.14055	.09119	-.00002	44.046	-.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	.00077	.00300	.22257	1.9197	.00784	24.665	.07651
Stddev	.00017	.00044	.00144	.01624	.0794	.00414	.178	.00017
%RSD	34.605	57.594	48.079	7.2986	4.1341	52.815	.72144	.22200
#1	.00066	.00116	.00292	.22095	1.8333	.01224	24.592	.07669
#2	.00046	.00029	.00448	.23956	1.9894	.00402	24.534	.07635
#3	.00033	.00086	.00160	.20719	1.9365	.00726	24.867	.07649
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00251	121.52	.00039	.04724	.00176	-.00202	.00114	5.9103
Stddev	.00040	.42	.00024	.00471	.00159	.00145	.00478	.0521
%RSD	16.119	.34732	62.165	9.9710	90.693	71.551	420.31	.88218
#1	.00228	121.85	.00064	.05267	.00005	-.00053	.00361	5.8778
#2	.00297	121.67	.00016	.04481	.00202	-.00342	.00418	5.9704
#3	.00227	121.04	.00037	.04425	.00320	-.00211	-.00437	5.8826
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115105 Acquired: 6/25/2016 14:35:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00058	.28912	-.00001	-.00113	.00539	.00821	.00159
Stddev	.00104	.00030	.00390	.00110	.00033	.00010	.00078
%RSD	180.18	.10436	27479.	97.280	6.0662	1.2671	48.813
#1	.00169	.28878	.00322	.00014	.00577	.00817	.00171
#2	.00041	.28919	.00108	-.00182	.00517	.00833	.00230
#3	-.00037	.28937	-.00434	-.00172	.00524	.00813	.00076
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9591.1	58078.	3001.7
Stddev	50.0	111.	7.0
%RSD	.52084	.19101	.23475
#1	9648.1	57963.	2993.9
#2	9554.9	58185.	3007.6
#3	9570.3	58085.	3003.7

Approved: June 27, 2016

Sample Name: L1606115401 Acquired: 6/25/2016 14:39:13 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00028**	**.06369**	**.00211**	**1.6713**	**.01926**	**.00010**	F **452.03**
Stddev	.00322	.00656	.00351	.0005	.00118	.00010	.89
%RSD	1157.9	10.295	166.24	.02810	6.1016	104.76	.19751
#1	.00114	.06870	-.00193	1.6717	.01795	.00008	451.94
#2	.00198	.05626	.00441	1.6715	.02023	.00000	452.97
#3	-.00396	.06610	.00386	1.6708	.01959	.00021	451.19
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail
High Limit							270.00
Low Limit							-.10000

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00042**	**.00175**	**.00336**	**.00358**	**.08657**	**4.0470**	**.10636**
Stddev	.00024	.00006	.00098	.00171	.01546	.0608	.00217
%RSD	56.333	3.4162	29.186	47.723	17.854	1.5017	2.0382
#1	.00015	.00180	.00230	.00410	.09591	4.0001	.10764
#2	.00053	.00168	.00424	.00167	.09508	4.1156	.10386
#3	.00059	.00177	.00354	.00498	.06873	4.0251	.10759
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**104.45**	**.25861**	**.02316**	F **889.97**	**.00046**	**.03619**	**.00320**
Stddev	.32	.00175	.00063	8.94	.00094	.00407	.00206
%RSD	.30573	.67823	2.7369	1.0045	203.65	11.246	64.428
#1	104.40	.26014	.02389	897.40	-.00019	.03795	.00082
#2	104.79	.25898	.02276	892.47	.00155	.03153	.00442
#3	104.16	.25670	.02283	880.05	.00003	.03907	.00435
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 27, 2016

Sample Name: L1606115401 Acquired: 6/25/2016 14:39:13 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-02

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**.04556**	**5.0652**	**-.00092**	**8.3910**	F **-.05563**	**-.00220**
Stddev	.00340	.00461	.0400	.00030	.0291	.00259	.00257
%RSD	2618.0	10.119	.78890	32.506	.34723	4.6498	116.61
#1	-.00398	.04208	5.1054	-.00114	8.4061	-.05807	-.00507
#2	.00246	.05079	5.0645	-.00058	8.4095	-.05590	-.00010
#3	.00113	.04380	5.0255	-.00104	8.3574	-.05292	-.00145
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.01649**	**.00488**	**.00303**
Stddev	.00056	.00014	.00088
%RSD	3.4018	2.9544	29.046
#1	.01604	.00505	.00374
#2	.01712	.00480	.00205
#3	.01632	.00480	.00329
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**8680.3**	**53198.**	**2985.6**
Stddev	180.6	1315.	22.8
%RSD	2.0805	2.4712	.76305
#1	8472.3	54211.	3004.6
#2	8770.9	53671.	2960.4
#3	8797.6	51712.	2991.9

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 14:43:13 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.41555	10.459	.41251	.50877	1.0374	.05191	10.456
Stddev	.00198	.038	.00049	.00412	.0216	.00011	.214
%RSD	.47730	.36575	.11809	.80911	2.0778	.21811	2.0454
#1	.41688	10.435	.41304	.50472	1.0594	.05178	10.677
#2	.41327	10.438	.41238	.51295	1.0163	.05199	10.250
#3	.41650	10.503	.41210	.50863	1.0364	.05195	10.439
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05257	.20726	.51590	.53004	4.1313	50.651	1.0395
Stddev	.00009	.00077	.00185	.00200	.0874	1.016	.0162
%RSD	.16567	.37212	.35803	.37798	2.1147	2.0061	1.5578
#1	.05256	.20724	.51397	.52980	4.1823	51.735	1.0578
#2	.05249	.20650	.51606	.52817	4.0304	49.720	1.0270
#3	.05266	.20804	.51766	.53216	4.1811	50.499	1.0337
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	10.150	.52278	1.0570	51.369	.52494	10.305	.52104
Stddev	.208	.00621	.0043	.990	.00089	.023	.00138
%RSD	2.0459	1.1888	.41022	1.9281	.16882	.22613	.26480
#1	10.382	.52905	1.0620	52.419	.52481	10.311	.51949
#2	10.083	.51662	1.0546	50.452	.52412	10.279	.52151
#3	9.9835	.52267	1.0544	51.234	.52588	10.325	.52213
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 14:43:13 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2435**	**.41340**	**5.3297**	**1.0365**	**1.0404**	**1.0490**	**.52867**
Stddev	.0031	.00239	.0078	.0027	.0194	.0184	.00366
%RSD	.25127	.57838	.14640	.26182	1.8676	1.7498	.69286
#1	1.2461	.41607	5.3340	1.0379	1.0597	1.0678	.52786
#2	1.2400	.41146	5.3207	1.0334	1.0208	1.0311	.53267
#3	1.2443	.41266	5.3345	1.0383	1.0408	1.0483	.52548
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0507**	**1.0444**	F **3.0477**
Stddev	.0007	.0005	.0535
%RSD	.06848	.05207	1.7552
#1	1.0503	1.0446	3.1080
#2	1.0502	1.0438	3.0059
#3	1.0515	1.0449	3.0292
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9539.3**	**57751.**	**2940.1**
Stddev	31.6	140.	48.7
%RSD	.33120	.24295	1.6549
#1	9524.0	57816.	2886.9
#2	9575.6	57847.	2982.3
#3	9518.3	57590.	2951.2

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 14:46:39 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00165	.00284	.00176	.00388	.00113	.00012	-.05362	.00014
Stddev	.00125	.01178	.00236	.00154	.00084	.00005	.04538	.00017
%RSD	75.692	415.29	134.23	39.663	74.227	41.354	84.617	118.22
#1	.00309	-.01077	.00340	.00450	.00053	.00007	-.00140	.00006
#2	.00089	.00942	.00281	.00212	.00078	.00011	-.07612	.00034
#3	.00097	.00986	-.00094	.00500	.00209	.00017	-.08336	.00003
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00024	.00081	-.00057	.03333	.05011	.00030	.06903	-.00046
Stddev	.00028	.00137	.00241	.00857	.04565	.00118	.08939	.00080
%RSD	113.95	168.04	420.16	25.719	91.103	392.02	129.50	172.20
#1	.00053	-.00076	.00045	.02973	.09790	.00058	.12386	.00033
#2	.00023	.00164	-.00333	.02715	.04546	-.00100	-.03412	-.00045
#3	-.00003	.00157	.00116	.04312	.00696	.00132	.11734	-.00127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00329	.09636	-.00015	.00820	-.00027	.00540	.00512	-.00542
Stddev	.00036	.02670	.00052	.00425	.00262	.00178	.00316	.00074
%RSD	10.874	27.706	348.06	51.818	968.89	32.864	61.650	13.619
#1	.00316	.11195	.00032	.01302	.00232	.00377	.00634	-.00562
#2	.00301	.11159	-.00071	.00662	-.00021	.00729	.00749	-.00604
#3	.00369	.06553	-.00007	.00498	-.00292	.00514	.00154	-.00460
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 14:46:39 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00030	.00084	-.00314	-.00152	.00124	-.00002	.00422
Stddev	.00043	.00035	.00315	.00163	.00162	.00008	.00071
%RSD	142.14	41.948	100.51	107.27	131.28	387.02	16.743
#1	.00054	.00124	-.00369	-.00248	.00007	.00002	.00481
#2	.00057	.00059	.00026	.00036	.00055	-.00012	.00343
#3	-.00019	.00069	-.00597	-.00244	.00309	.00004	.00441
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9509.0	57769.	2936.8
Stddev	19.9	100.	17.6
%RSD	.20940	.17266	.59912
#1	9496.2	57883.	2947.5
#2	9532.0	57728.	2946.4
#3	9498.9	57697.	2916.5

Approved: June 27, 2016

Sample Name: L1606115402S Acquired: 6/25/2016 14:50:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.20060**	**4.9650**	**.19517**	**2.5768**	**.50183**	**.02406**	F **444.38**
Stddev	.00064	.0130	.00068	.0061	.00191	.00013	.71
%RSD	.32075	.26173	.35003	.23805	.38088	.54060	.15891
#1	.20131	4.9565	.19473	2.5781	.50380	.02409	444.63
#2	.20005	4.9586	.19596	2.5701	.50170	.02392	443.58
#3	.20045	4.9800	.19484	2.5821	.49998	.02418	444.92
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail
High Limit							270.00
Low Limit							-.10000

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02442**	**.09295**	**.23551**	**.23263**	**1.9191**	**28.985**	**.61166**
Stddev	.00012	.00006	.00093	.00115	.0123	.093	.00400
%RSD	.49172	.06109	.39479	.49604	.64319	.32038	.65326
#1	.02431	.09291	.23631	.23296	1.9231	29.086	.60738
#2	.02455	.09302	.23573	.23359	1.9290	28.966	.61529
#3	.02440	.09293	.23449	.23135	1.9053	28.904	.61230
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**105.95**	**.47454**	**.51043**	F **884.75**	**.22811**	**4.9353**	**.22750**
Stddev	.25	.00567	.00309	9.80	.00122	.0272	.00524
%RSD	.23467	1.1953	.60519	1.1073	.53281	.55168	2.3044
#1	105.68	.47627	.51271	882.43	.22677	4.9309	.23221
#2	106.17	.47914	.51168	876.32	.22915	4.9645	.22843
#3	105.99	.46820	.50692	895.49	.22841	4.9105	.22185
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 27, 2016

Sample Name: L1606115402S Acquired: 6/25/2016 14:50:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-07

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.58292**	**.23516**	**7.5230**	**.45764**	**8.4638**	**.43138**	**.22024**
Stddev	.00661	.00590	.0386	.00329	.0240	.00658	.00196
%RSD	1.1347	2.5090	.51250	.71986	.28403	1.5254	.88967
#1	.58964	.23313	7.5444	.46024	8.4871	.43835	.22044
#2	.58270	.23054	7.5461	.45874	8.4651	.43053	.22209
#3	.57642	.24181	7.4785	.45394	8.4391	.42527	.21819
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.50974**	**.45745**	**.02017**
Stddev	.00182	.00152	.00279
%RSD	.35750	.33182	13.807
#1	.50798	.45815	.01724
#2	.50964	.45849	.02049
#3	.51162	.45571	.02279
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**8784.0**	**53813.**	**2939.5**
Stddev	58.9	564.	26.0
%RSD	.67031	1.0483	.88591
#1	8819.5	54202.	2951.8
#2	8816.5	54070.	2957.2
#3	8716.1	53166.	2909.6

Approved: June 27, 2016

Sample Name: L1606115403SD Acquired: 6/25/2016 14:54:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19884**	**5.0097**	**.20311**	**2.5960**	**.50146**	**.02443**	F **447.43**
Stddev	.00024	.0049	.00234	.0007	.00093	.00023	1.15
%RSD	.12160	.09827	1.1515	.02603	.18621	.94697	.25628
#1	.19893	5.0097	.20547	2.5958	.50254	.02464	447.39
#2	.19857	5.0047	.20308	2.5967	.50101	.02419	446.31
#3	.19903	5.0146	.20079	2.5954	.50084	.02448	448.60
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail
High Limit							270.00
Low Limit							-.10000

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02483**	**.09451**	**.23821**	**.23915**	**1.9535**	**28.902**	**.60873**
Stddev	.00032	.00036	.00123	.00189	.0137	.108	.00282
%RSD	1.3026	.38369	.51502	.78901	.69899	.37351	.46252
#1	.02509	.09492	.23941	.23914	1.9479	29.020	.60854
#2	.02447	.09421	.23696	.23726	1.9691	28.808	.61163
#3	.02493	.09441	.23826	.24103	1.9435	28.877	.60601
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**106.60**	**.48110**	**.52231**	F **881.76**	**.23168**	**5.0586**	**.22685**
Stddev	.66	.00483	.00325	3.70	.00136	.0426	.00410
%RSD	.61507	1.0048	.62269	.41950	.58652	.84170	1.8079
#1	107.35	.47570	.52433	877.70	.23164	5.0964	.23130
#2	106.14	.48502	.51856	884.94	.23034	5.0125	.22604
#3	106.32	.48257	.52404	882.64	.23305	5.0670	.22322
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 27, 2016

Sample Name: L1606115403SD Acquired: 6/25/2016 14:54:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-08

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.59871**	**.24615**	**7.5687**	**.46699**	**8.5686**	**.42893**	**.22250**
Stddev	.00609	.00268	.0516	.00318	.0150	.00524	.00079
%RSD	1.0180	1.0870	.68153	.68125	.17550	1.2224	.35602
#1	.60575	.24920	7.6059	.46994	8.5810	.42717	.22334
#2	.59510	.24505	7.5098	.46362	8.5519	.42478	.22177
#3	.59529	.24419	7.5903	.46740	8.5730	.43482	.22237
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.51554**	**.46791**	**.00590**
Stddev	.00333	.00294	.00185
%RSD	.64552	.62860	31.408
#1	.51938	.47002	.00382
#2	.51356	.46455	.00651
#3	.51367	.46915	.00737
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**8900.3**	**52138.**	**2849.7**
Stddev	61.9	226.	11.1
%RSD	.69538	.43440	.38823
#1	8858.0	51879.	2838.3
#2	8971.3	52295.	2860.4
#3	8871.6	52242.	2850.4

Approved: June 27, 2016

Sample Name: L1606115404 Acquired: 6/25/2016 14:58:09 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.00639	-.00036	.01270	.00241	.00013	.28017	.00009
Stddev	.00146	.00656	.00244	.00114	.00043	.00008	.03044	.00010
%RSD	248.07	102.69	686.61	9.0101	18.005	61.828	10.864	104.11
#1	.00089	-.00046	-.00313	.01156	.00227	.00022	.24651	.00012
#2	.00188	.01261	.00061	.01385	.00206	.00012	.28824	-.00001
#3	-.00100	.00700	.00146	.01270	.00289	.00005	.30576	.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	.01305	.00240	.07792	.04064	-.00228	.12469	.00253
Stddev	.00039	.00070	.00120	.00694	.09353	.00295	.18212	.00145
%RSD	747.32	5.3526	50.113	8.9064	230.14	129.38	146.06	57.512
#1	.00045	.01376	.00301	.08066	.00523	-.00157	.14106	.00090
#2	-.00033	.01305	.00317	.08306	.14671	-.00552	.29807	.00300
#3	.00004	.01236	.00101	.07002	-.03002	.00025	-.06507	.00369
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00228	.75291	.00807	.00690	.00300	-.00255	.00146	.13135
Stddev	.00024	.01638	.00065	.00315	.00148	.00424	.00328	.00151
%RSD	10.641	2.1753	8.1061	45.563	49.471	166.69	225.01	1.1485
#1	.00245	.73694	.00868	.00525	.00160	-.00242	-.00127	.13296
#2	.00238	.75210	.00738	.01053	.00455	-.00685	.00055	.13112
#3	.00200	.76967	.00814	.00493	.00285	.00163	.00509	.12997
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115404 Acquired: 6/25/2016 14:58:09 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00160	.00282	.00431	-.00324	.00105	.00371	.00033
Stddev	.00036	.00041	.00156	.00193	.00089	.00012	.00066
%RSD	22.302	14.434	36.136	59.618	84.281	3.3596	201.18
#1	.00136	.00295	.00420	-.00210	.00202	.00385	-.00028
#2	.00201	.00236	.00592	-.00547	.00085	.00361	.00023
#3	.00143	.00314	.00281	-.00214	.00028	.00369	.00104
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9713.7	59908.	3002.7
Stddev	32.6	164.	34.0
%RSD	.33561	.27353	1.1337
#1	9740.1	60011.	2965.4
#2	9677.2	59995.	3032.2
#3	9723.7	59719.	3010.3

Approved: June 27, 2016

Sample Name: L1606115501 Acquired: 6/25/2016 15:02:07 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00226	.00780	.00401	.26646	.17918	.00000	45.474	-.00004
Stddev	.00287	.00301	.00103	.00133	.00076	.00004	.222	.00008
%RSD	126.68	38.602	25.778	.50014	.42327	919.30	.48843	228.57
#1	.00509	.01128	.00462	.26750	.17974	.00002	45.718	.00006
#2	.00233	.00600	.00459	.26694	.17949	-.00004	45.422	-.00008
#3	-.00064	.00613	.00281	.26496	.17832	.00004	45.283	-.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.00088	.00134	.05774	1.5654	.00884	27.038	1.6699
Stddev	.00027	.00028	.00194	.01943	.1007	.00205	.083	.0027
%RSD	44.880	31.682	144.85	33.650	6.4312	23.214	.30796	.16065
#1	.00083	.00056	-.00088	.03531	1.4495	.00687	26.967	1.6730
#2	.00063	.00109	.00219	.06938	1.6158	.01096	27.130	1.6684
#3	.00031	.00099	.00272	.06852	1.6309	.00869	27.018	1.6684
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00414	74.521	.00232	.02280	.00145	-.00333	.00418	6.9589
Stddev	.00012	.362	.00063	.00622	.00648	.00099	.00335	.0172
%RSD	2.8693	.48589	27.089	27.292	447.38	29.598	80.140	.24675
#1	.00419	74.939	.00163	.02521	-.00229	-.00322	.00046	6.9755
#2	.00400	74.314	.00245	.02746	-.00229	-.00437	.00510	6.9601
#3	.00421	74.311	.00286	.01573	.00893	-.00240	.00696	6.9412
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115501 Acquired: 6/25/2016 15:02:07 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00017**	**.31080**	**-.00408**	**-.00246**	**.00074**	**.00517**	**.00097**
Stddev	.00101	.00127	.01220	.00176	.00032	.00009	.00026
%RSD	589.50	.40720	299.39	71.557	43.195	1.7947	26.599
#1	.00067	.31223	.00838	-.00158	.00040	.00528	.00122
#2	-.00129	.30982	-.01601	-.00449	.00079	.00510	.00071
#3	.00011	.31036	-.00459	-.00131	.00103	.00514	.00097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9554.2**	**58064.**	**2987.7**
Stddev	24.6	152.	17.9
%RSD	.25722	.26211	.59940
#1	9527.5	57892.	2971.4
#2	9576.0	58182.	2984.9
#3	9559.0	58118.	3006.9

Approved: June 27, 2016

Sample Name: L1606115502 Acquired: 6/25/2016 15:05:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00077	.00104	.00250	.19417	.33327	.00001	49.969	.00011
Stddev	.00225	.00260	.00021	.00180	.00208	.00007	.144	.00004
%RSD	293.51	250.20	8.2634	.92950	.62356	535.11	.28842	36.458
#1	.00053	-.00100	.00271	.19484	.33088	.00006	50.109	.00015
#2	-.00336	.00396	.00251	.19555	.33470	.00004	49.821	.00012
#3	.00053	.00015	.00229	.19213	.33422	-.00007	49.978	.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00029	.00039	.00161	4.4932	.30074	.00699	33.094	.46968
Stddev	.00024	.00046	.00055	.0210	.07361	.00374	.139	.00201
%RSD	82.593	118.92	34.456	.46819	24.475	53.562	.42105	.42764
#1	.00002	-.00012	.00218	4.4716	.35097	.00318	33.254	.46985
#2	.00046	.00051	.00108	4.5136	.21625	.00712	33.021	.47160
#3	.00039	.00078	.00156	4.4944	.33501	.01067	33.006	.46759
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00134	56.528	.00048	.02851	.00540	-.00461	.00421	7.9465
Stddev	.00013	.068	.00114	.00462	.00195	.00412	.00583	.0128
%RSD	9.8009	.11946	234.80	16.220	36.120	89.402	138.37	.16106
#1	.00123	56.559	.00179	.03249	.00375	-.00852	-.00246	7.9518
#2	.00130	56.451	-.00006	.02962	.00489	-.00500	.00681	7.9559
#3	.00149	56.575	-.00027	.02344	.00755	-.00031	.00829	7.9320
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115502 Acquired: 6/25/2016 15:05:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00013**	**.41936**	**-.00337**	**-.00247**	**.00035**	**.00337**	**.00091**
Stddev	.00022	.00050	.00295	.00045	.00041	.00006	.00160
%RSD	168.21	.11997	87.675	18.352	119.11	1.8649	175.07
#1	.00016	.41971	-.00131	-.00286	-.00012	.00341	-.00093
#2	.00033	.41958	-.00675	-.00259	.00050	.00330	.00173
#3	-.00010	.41878	-.00204	-.00197	.00066	.00341	.00194
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9627.3**	**58565.**	**2988.6**
Stddev	17.2	39.	5.1
%RSD	.17881	.06667	.17070
#1	9625.0	58521.	2993.1
#2	9645.6	58579.	2989.5
#3	9611.4	58596.	2983.0

Approved: June 27, 2016

Sample Name: L1606115503 Acquired: 6/25/2016 15:09:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00133	.01981	.00070	.15980	.12508	.00001	47.416	.00012
Stddev	.00149	.00690	.00351	.00313	.00043	.00010	.201	.00026
%RSD	112.43	34.815	504.28	1.9589	.34619	838.85	.42403	214.19
#1	.00108	.02777	-.00104	.15681	.12534	.00005	47.373	.00032
#2	-.00003	.01583	-.00161	.16306	.12532	.00009	47.635	.00022
#3	.00292	.01581	.00474	.15953	.12458	-.00010	47.240	-.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00011	.00032	.00351	1.3413	.87483	.00900	17.381	.47700
Stddev	.00046	.00045	.00082	.0140	.08308	.00172	.089	.00081
%RSD	429.81	139.46	23.247	1.0398	9.4971	19.124	.51021	.16934
#1	.00045	.00052	.00286	1.3297	.89461	.00799	17.319	.47725
#2	-.00041	.00065	.00443	1.3568	.78364	.01099	17.483	.47609
#3	.00028	-.00019	.00325	1.3373	.94624	.00803	17.342	.47764
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00026	48.369	.00082	.10919	.00167	-.00062	.00179	5.8799
Stddev	.00025	.203	.00140	.00660	.00170	.00395	.00827	.0280
%RSD	97.415	.41932	171.45	6.0459	101.94	641.48	462.74	.47549
#1	-.00003	48.343	.00233	.10833	.00101	-.00514	.00952	5.8933
#2	-.00053	48.584	.00055	.10306	.00039	.00119	.00277	5.8478
#3	-.00022	48.181	-.00043	.11618	.00360	.00211	-.00693	5.8986
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115503 Acquired: 6/25/2016 15:09:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00019**	**.26600**	**-.00537**	**-.00338**	**.00017**	**.00914**	**.00020**
Stddev	.00067	.00095	.00241	.00310	.00116	.00022	.00114
%RSD	345.33	.35889	44.987	91.459	671.24	2.4286	566.35
#1	-.00057	.26556	-.00315	-.00551	-.00106	.00889	.00099
#2	.00050	.26710	-.00794	.00017	.00033	.00931	-.00111
#3	.00065	.26535	-.00501	-.00481	.00125	.00922	.00072
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9587.2**	**58263.**	**2972.1**
Stddev	19.6	279.	4.9
%RSD	.20409	.47925	.16604
#1	9598.2	58004.	2969.9
#2	9598.7	58226.	2968.6
#3	9564.6	58559.	2977.8

Approved: June 27, 2016

Sample Name: L1606117801 Acquired: 6/25/2016 15:13:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00123	.01562	.00788	.04060	.02600	.00012	2.4592	.00007
Stddev	.00244	.00616	.00118	.00330	.00129	.00005	.0606	.00006
%RSD	197.99	39.461	15.044	8.1358	4.9545	44.901	2.4620	79.703
#1	.00387	.02096	.00725	.04075	.02646	.00010	2.5149	.00002
#2	-.00097	.00887	.00713	.04382	.02700	.00008	2.4680	.00013
#3	.00081	.01703	.00924	.03722	.02455	.00018	2.3948	.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00003	.00074	.00609	.02685	.43656	.00432	.57708	-.00111
Stddev	.00033	.00144	.00018	.01358	.05030	.00166	.05180	.00137
%RSD	1012.2	194.14	2.9070	50.575	11.522	38.411	8.9754	123.48
#1	-.00025	.00070	.00629	.04250	.40355	.00528	.52291	-.00216
#2	.00039	.00219	.00602	.01809	.49445	.00241	.62611	-.00159
#3	-.00004	-.00068	.00595	.01997	.41167	.00528	.58223	.00044
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00299	95.163	-.00070	.01050	.00222	-.00635	.00641	3.4783
Stddev	.00029	.186	.00058	.00517	.00074	.00345	.00307	.0048
%RSD	9.6245	.19542	82.717	49.206	33.276	54.334	47.947	.13892
#1	.00325	95.230	-.00082	.01297	.00142	-.01032	.00356	3.4792
#2	.00268	95.306	-.00121	.00456	.00236	-.00408	.00600	3.4827
#3	.00303	94.953	-.00007	.01396	.00287	-.00464	.00966	3.4731
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606117801 Acquired: 6/25/2016 15:13:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00037	.11699	.00044	-.00188	.00105	.00210	.00054
Stddev	.00103	.00010	.00531	.00109	.00009	.00007	.00166
%RSD	275.08	.08958	1212.8	57.943	8.6567	3.2366	310.87
#1	.00154	.11703	-.00019	-.00063	.00114	.00212	-.00081
#2	-.00042	.11707	-.00453	-.00235	.00105	.00215	.00002
#3	-.00000	.11687	.00604	-.00264	.00096	.00202	.00240
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9648.4	58829.	2994.3
Stddev	5.2	37.	6.7
%RSD	.05359	.06360	.22320
#1	9653.7	58809.	2992.7
#2	9643.3	58806.	2988.6
#3	9648.3	58873.	3001.7

Approved: June 27, 2016

Sample Name: L1606122701 Acquired: 6/25/2016 15:17:43 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00116	.09783	-.00055	.00242	.01519	.00006	7.7326	-.00002
Stddev	.00092	.00377	.00239	.00147	.00198	.00002	.1171	.00010
%RSD	79.355	3.8537	437.80	60.740	13.003	38.615	1.5138	459.39
#1	-.00047	.10163	-.00294	.00351	.01401	.00007	7.7569	.00006
#2	-.00220	.09409	.00184	.00075	.01747	.00007	7.8356	-.00013
#3	-.00081	.09776	-.00054	.00300	.01410	.00003	7.6053	.00000
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00027	.00227	.00167	.03463	.33602	-.00038	-.01109	-.00127
Stddev	.00016	.00059	.00051	.00142	.06178	.00154	.10836	.00134
%RSD	58.524	25.933	30.520	4.0976	18.384	404.83	977.09	105.32
#1	.00014	.00159	.00133	.03617	.30512	-.00005	-.09610	-.00256
#2	.00023	.00261	.00225	.03337	.29580	-.00206	.11092	-.00136
#3	.00045	.00260	.00142	.03436	.40715	.00097	-.04809	.00011
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00015	2.3350	.00044	.00731	.00430	-.00426	.00383	1.1328
Stddev	.00020	.0307	.00076	.00233	.00100	.00159	.00576	.0032
%RSD	132.69	1.3153	170.74	31.908	23.273	37.398	150.44	.27905
#1	.00005	2.3000	.00116	.00516	.00531	-.00538	.00306	1.1309
#2	-.00036	2.3574	.00052	.00700	.00430	-.00496	-.00151	1.1310
#3	-.00016	2.3476	-.00035	.00979	.00330	-.00244	.00992	1.1364
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606122701 Acquired: 6/25/2016 15:17:43 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	.25125	.00414	.00180	.00364	.00056	.00136
Stddev	.00018	.00103	.01324	.00145	.00074	.00007	.00130
%RSD	364.81	.41002	320.05	80.159	20.271	13.218	96.223
#1	.00011	.25131	-.01092	.00207	.00450	.00061	.00115
#2	.00019	.25224	.00936	.00024	.00328	.00047	.00016
#3	-.00016	.25018	.01397	.00310	.00316	.00058	.00275
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9969.0	60494.	3048.9
Stddev	15.1	129.	13.3
%RSD	.15159	.21283	.43564
#1	9969.3	60641.	3055.8
#2	9983.9	60438.	3033.6
#3	9953.7	60402.	3057.3

Approved: June 27, 2016

Sample Name: L1606122706 Acquired: 6/25/2016 15:21:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00063**	**.00567**	**.00083**	**.00123**	**.02806**	**.00015**	**13.472**	**-.00038**
Stddev	.00250	.00554	.00101	.00202	.00007	.00007	.038	.00019
%RSD	395.00	97.673	121.57	163.43	.25882	47.682	.28032	49.732
#1	.00247	-.00071	.00042	-.00047	.02810	.00007	13.487	-.00023
#2	.00165	.00924	.00199	.00346	.02810	.00021	13.429	-.00032
#3	-.00222	.00849	.00010	.00072	.02797	.00018	13.501	-.00060
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00101**	**.00061**	**.00157**	**.10367**	**.20316**	**.00472**	**8.1185**	**.10636**
Stddev	.00008	.00062	.00064	.01594	.00836	.00156	.1274	.00158
%RSD	8.0012	102.28	40.873	15.374	4.1155	32.981	1.5689	1.4897
#1	.00092	-.00009	.00190	.11049	.21157	.00610	7.9808	.10550
#2	.00107	.00111	.00198	.11507	.19485	.00503	8.2322	.10539
#3	.00104	.00079	.00083	.08546	.20306	.00303	8.1425	.10819
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00044**	**33.974**	**.00004**	**.00429**	**-.00079**	**-.00333**	**.00187**	**.74797**
Stddev	.00043	.239	.00011	.00671	.00165	.00196	.00470	.01610
%RSD	97.916	.70374	251.40	156.52	208.37	58.883	251.58	2.1527
#1	-.00064	33.727	-.00007	.01191	-.00018	-.00196	-.00208	.73053
#2	-.00073	33.993	.00015	-.00074	.00047	-.00558	.00061	.75112
#3	.00005	34.204	.00006	.00169	-.00266	-.00246	.00707	.76226
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606122706 Acquired: 6/25/2016 15:21:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00014**	**.30857**	**.00233**	**-.00259**	**.00019**	**.00158**	**.00089**
Stddev	.00012	.00270	.00311	.00390	.00060	.00013	.00062
%RSD	87.894	.87470	133.51	150.50	319.58	8.2216	69.559
#1	-.00001	.30602	.00364	-.00319	-.00046	.00143	.00122
#2	-.00015	.30828	-.00122	.00157	.00073	.00163	.00127
#3	-.00026	.31139	.00457	-.00615	.00030	.00167	.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9924.8**	**59991.**	**2995.5**
Stddev	18.6	172.	10.6
%RSD	.18782	.28732	.35313
#1	9924.0	59808.	2993.5
#2	9943.8	60150.	3007.0
#3	9906.6	60016.	2986.1

Approved: June 27, 2016

Sample Name: L1606127902 Acquired: 6/25/2016 15:25:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00298**	**.77289**	**.00033**	**.00803**	**.01053**	**.00016**	**9.3649**	**-.00023**
Stddev	.00188	.01227	.00104	.00267	.00091	.00007	.0910	.00007
%RSD	63.015	1.5872	311.80	33.213	8.6722	42.111	.97197	32.230
#1	.00105	.75975	.00114	.01109	.01057	.00009	9.3786	-.00026
#2	.00309	.78404	.00069	.00619	.01142	.00023	9.4483	-.00028
#3	.00480	.77489	-.00084	.00682	.00960	.00016	9.2678	-.00014
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00067**	**.00209**	**.00225**	**1.1046**	**.20587**	**.00183**	**3.8329**	**.12468**
Stddev	.00025	.00124	.00151	.0075	.05413	.00323	.0374	.00185
%RSD	36.537	59.143	66.907	.67811	26.294	176.78	.97568	1.4875
#1	.00057	.00084	.00284	1.1048	.18555	.00515	3.8178	.12286
#2	.00096	.00331	.00337	1.1121	.26722	.00166	3.8755	.12657
#3	.00050	.00212	.00054	1.0971	.16483	-.00131	3.8054	.12461
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00011**	**.91553**	**.00104**	**.04284**	**.00087**	**-.00434**	**.00416**	**1.4218**
Stddev	.00049	.02558	.00097	.00374	.00128	.00145	.00383	.0304
%RSD	448.71	2.7944	92.896	8.7405	147.22	33.355	92.023	2.1368
#1	.00043	.88763	.00204	.04714	.00040	-.00289	.00799	1.3901
#2	-.00054	.92105	.00011	.04033	-.00011	-.00435	.00416	1.4246
#3	-.00022	.93790	.00097	.04105	.00232	-.00579	.00033	1.4507
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127902 Acquired: 6/25/2016 15:25:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00026	.04596	.00940	.00124	.00257	.00907	.00122
Stddev	.00038	.00038	.00559	.00152	.00046	.00027	.00089
%RSD	147.88	.81748	59.487	122.54	17.946	3.0009	73.195
#1	.00017	.04588	.00305	-.00019	.00219	.00893	.00224
#2	-.00038	.04637	.01155	.00284	.00243	.00890	.00056
#3	-.00056	.04564	.01360	.00108	.00308	.00938	.00087
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10039.	61079.	3063.3
Stddev	16.	87.	16.8
%RSD	.15707	.14311	.54820
#1	10022.	60979.	3046.2
#2	10041.	61122.	3064.0
#3	10054.	61137.	3079.8

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 15:29:11 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.41474	10.480	.41122	.51129	1.0445	.05223	10.384
Stddev	.00157	.008	.00187	.00319	.0019	.00016	.024
%RSD	.37799	.07401	.45408	.62455	.18188	.31144	.22798
#1	.41325	10.484	.41182	.51035	1.0465	.05219	10.387
#2	.41461	10.484	.41272	.50866	1.0441	.05241	10.359
#3	.41637	10.471	.40913	.51484	1.0428	.05209	10.407
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05244	.20643	.51777	.52807	4.1528	50.626	1.0452
Stddev	.00019	.00026	.00163	.00025	.0357	.257	.0080
%RSD	.36313	.12673	.31564	.04665	.85968	.50669	.76446
#1	.05244	.20669	.51592	.52825	4.1190	50.704	1.0431
#2	.05226	.20644	.51838	.52817	4.1901	50.835	1.0540
#3	.05264	.20617	.51902	.52779	4.1492	50.340	1.0385
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.9629	.52245	1.0531	51.247	.52506	10.312	.52216
Stddev	.0274	.00433	.0047	.134	.00127	.013	.00461
%RSD	.27495	.82847	.44799	.26095	.24264	.12919	.88203
#1	9.9944	.51946	1.0578	51.179	.52653	10.320	.51773
#2	9.9492	.52742	1.0532	51.401	.52432	10.296	.52692
#3	9.9451	.52049	1.0484	51.161	.52433	10.319	.52182
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 15:29:11 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2356**	**.41239**	**5.3381**	**1.0346**	**1.0460**	**1.0497**	**.53160**
Stddev	.0017	.00642	.0057	.0013	.0038	.0052	.00142
%RSD	.13528	1.5556	.10675	.12483	.36598	.49394	.26637
#1	1.2362	.41861	5.3408	1.0357	1.0443	1.0445	.53322
#2	1.2337	.40580	5.3420	1.0349	1.0504	1.0497	.53096
#3	1.2369	.41277	5.3316	1.0332	1.0433	1.0549	.53062
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0558**	**1.0420**	F **3.0350**
Stddev	.0013	.0007	.0090
%RSD	.12612	.06683	.29605
#1	1.0548	1.0423	3.0277
#2	1.0553	1.0412	3.0451
#3	1.0573	1.0426	3.0323
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9586.8**	**58077.**	**2975.5**
Stddev	6.0	176.	6.3
%RSD	.06238	.30386	.21060
#1	9587.3	57874.	2969.1
#2	9592.5	58193.	2975.6
#3	9580.6	58165.	2981.7

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 15:32:48 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00000**	**.00002**	**-.00010**	**.00491**	**.00210**	**.00011**	**.00565**	**-.00003**
Stddev	.00087	.00719	.00078	.00229	.00030	.00005	.00049	.00029
%RSD	26249.	33697.	808.92	46.626	14.233	46.346	8.5911	909.94
#1	.00025	.00801	-.00086	.00753	.00176	.00017	.00542	-.00000
#2	.00073	-.00593	.00069	.00390	.00223	.00007	.00533	-.00034
#3	-.00097	-.00201	-.00012	.00330	.00231	.00010	.00621	.00024
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00020**	**.00011**	**.00064**	**.03500**	**.03424**	**-.00179**	**.02594**	**-.00052**
Stddev	.00009	.00046	.00051	.01471	.03973	.00240	.07741	.00156
%RSD	43.655	413.87	80.028	42.019	116.03	134.20	298.41	302.41
#1	.00012	-.00018	.00036	.01804	-.01111	-.00218	.05944	-.00155
#2	.00020	.00064	.00123	.04417	.05092	-.00397	.08095	.00128
#3	.00029	-.00013	.00033	.04280	.06290	.00078	-.06257	-.00128
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00316**	**.04507**	**.00037**	**.00106**	**.00080**	**.00374**	**.00412**	**.00990**
Stddev	.00041	.02068	.00057	.00881	.00149	.00197	.00171	.00085
%RSD	13.022	45.885	154.95	828.64	185.93	52.502	41.416	8.5529
#1	.00281	.06772	-.00005	.00994	-.00040	.00294	.00217	.01067
#2	.00361	.02718	.00102	.00092	.00033	.00230	.00488	.01002
#3	.00306	.04031	.00014	-.00767	.00246	.00598	.00532	.00899
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 15:32:48 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00048**	**.00057**	**.00333**	**-.00225**	**.00169**	**-.00006**	**.00439**
Stddev	.00073	.00051	.00753	.00264	.00040	.00020	.00135
%RSD	152.10	89.662	226.48	117.22	23.766	306.69	30.851
#1	.00129	.00079	-.00329	-.00367	.00165	-.00017	.00283
#2	-.00014	-.00001	.01153	-.00387	.00130	-.00018	.00508
#3	.00029	.00092	.00174	.00079	.00210	.00016	.00526
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9759.3**	**59359.**	**3013.9**
Stddev	2.3	121.	1.6
%RSD	.02366	.20394	.05186
#1	9761.6	59292.	3015.4
#2	9759.3	59287.	3013.9
#3	9756.9	59499.	3012.3

Approved: June 27, 2016

Sample Name: L1606127903S Acquired: 6/25/2016 15:36:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01197	1.1036	.01087	.05712	.03304	.00134	9.9729	.00124
Stddev	.00119	.0059	.00187	.00108	.00116	.00008	.0996	.00005
%RSD	9.9297	.53710	17.218	1.8903	3.5218	6.1377	.99838	3.7414
#1	.01115	1.1017	.01246	.05619	.03414	.00126	10.019	.00125
#2	.01142	1.1103	.00881	.05688	.03182	.00135	9.8587	.00119
#3	.01333	1.0989	.01134	.05830	.03317	.00142	10.041	.00127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00560	.01397	.01462	1.0767	1.4208	.02780	4.2261	.14457
Stddev	.00014	.00108	.00214	.0241	.1192	.00407	.0440	.00290
%RSD	2.4102	7.7598	14.602	2.2369	8.3867	14.648	1.0401	2.0070
#1	.00574	.01498	.01254	1.0620	1.3125	.02932	4.2372	.14280
#2	.00559	.01283	.01680	1.0635	1.4014	.02319	4.1777	.14299
#3	.00547	.01411	.01453	1.1045	1.5484	.03090	4.2635	.14792
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02152	2.1329	.01322	.29583	.01584	.02588	.00942	1.7458
Stddev	.00098	.0418	.00158	.00727	.00355	.00244	.00136	.0125
%RSD	4.5315	1.9579	11.965	2.4574	22.390	9.4200	14.478	.71438
#1	.02231	2.1343	.01139	.30417	.01767	.02397	.01094	1.7523
#2	.02182	2.0905	.01413	.29245	.01809	.02862	.00829	1.7537
#3	.02043	2.1740	.01413	.29086	.01175	.02503	.00905	1.7314
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127903S Acquired: 6/25/2016 15:36:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02137**	**.07160**	**.03468**	**.01383**	**.02654**	**.02985**	**.00242**
Stddev	.00036	.00081	.00438	.00291	.00046	.00010	.00158
%RSD	1.6937	1.1317	12.615	21.070	1.7354	.32151	65.388
#1	.02178	.07099	.02963	.01549	.02613	.02986	.00141
#2	.02110	.07130	.03735	.01047	.02704	.02975	.00424
#3	.02123	.07252	.03706	.01554	.02644	.02994	.00160
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9975.1**	**60274.**	**3041.7**
Stddev	49.0	118.	22.7
%RSD	.49139	.19506	.74582
#1	9940.1	60350.	3061.9
#2	9954.2	60332.	3046.1
#3	10031.	60138.	3017.2

Approved: June 27, 2016

Sample Name: L1606127904SD Acquired: 6/25/2016 15:40:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.01103**	**1.1028**	**.01203**	**.05464**	**.03364**	**.00143**	**9.8642**	**.00121**
Stddev	.00186	.0090	.00180	.00231	.00086	.00004	.1305	.00006
%RSD	16.883	.81742	14.963	4.2286	2.5577	2.6326	1.3231	5.3301
#1	.00912	1.0960	.01267	.05217	.03303	.00146	9.7151	.00122
#2	.01113	1.1130	.00999	.05675	.03463	.00143	9.9198	.00114
#3	.01284	1.0994	.01342	.05499	.03327	.00139	9.9577	.00127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00590**	**.01420**	**.01484**	**1.0576**	**1.4598**	**.02493**	**4.2188**	**.14616**
Stddev	.00020	.00048	.00068	.0082	.1286	.00072	.1177	.00241
%RSD	3.3297	3.4140	4.5711	.77893	8.8104	2.9051	2.7895	1.6455
#1	.00602	.01364	.01562	1.0663	1.6059	.02466	4.0843	.14411
#2	.00567	.01444	.01454	1.0499	1.4099	.02439	4.2697	.14557
#3	.00601	.01451	.01437	1.0566	1.3636	.02576	4.3025	.14881
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02217**	**2.1040**	**.01418**	**.30006**	**.01556**	**.02408**	**.01453**	**1.7577**
Stddev	.00072	.0376	.00057	.00408	.00448	.00503	.00712	.0151
%RSD	3.2386	1.7891	4.0074	1.3597	28.796	20.896	49.023	.85907
#1	.02294	2.0606	.01353	.29698	.01280	.02854	.02252	1.7414
#2	.02203	2.1275	.01453	.30469	.02073	.02509	.00885	1.7604
#3	.02153	2.1239	.01449	.29851	.01315	.01862	.01221	1.7712
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127904SD Acquired: 6/25/2016 15:40:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02276**	**.07116**	**.03577**	**.01268**	**.02759**	**.03003**	**.00230**
Stddev	.00014	.00090	.00737	.00234	.00032	.00029	.00174
%RSD	.61715	1.2587	20.620	18.483	1.1623	.96154	75.576
#1	.02272	.07054	.03703	.01291	.02794	.02969	.00432
#2	.02292	.07076	.02784	.01490	.02752	.03021	.00132
#3	.02264	.07219	.04242	.01023	.02731	.03018	.00128
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10006.**	**60810.**	**3063.3**
Stddev	5.	192.	18.4
%RSD	.04992	.31626	.60150
#1	10009.	61010.	3082.6
#2	10009.	60627.	3061.4
#3	10000.	60793.	3045.9

Approved: June 27, 2016

Sample Name: L1606128901 Acquired: 6/25/2016 15:44:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00215	.00687	.00324	.00036	.25940	.00003	8.5369	.00005
Stddev	.00051	.01020	.00141	.00225	.00475	.00006	.1296	.00011
%RSD	23.677	148.57	43.630	628.60	1.8323	170.08	1.5176	227.30
#1	.00216	.01614	.00358	-.00039	.25471	.00001	8.3890	.00017
#2	.00265	.00852	.00169	.00289	.26421	.00010	8.6303	-.00003
#3	.00163	-.00406	.00445	-.00143	.25928	-.00000	8.5913	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00259	.00588	.00175	.08631	.04218	.00151	6.5072	.07542
Stddev	.00040	.00134	.00073	.02244	.04887	.00193	.2142	.00074
%RSD	15.316	22.711	41.571	26.002	115.87	127.81	3.2912	.97595
#1	.00214	.00436	.00096	.06469	.09555	-.00070	6.2873	.07459
#2	.00279	.00683	.00239	.10949	-.00039	.00286	6.5191	.07570
#3	.00285	.00647	.00190	.08474	.03137	.00237	6.7151	.07598
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	23.216	.03092	.00614	.00300	-.00111	.00640	.97050
Stddev	.00039	.461	.00078	.00363	.00087	.00404	.00265	.00247
%RSD	280.63	1.9852	2.5231	59.153	29.021	362.80	41.403	.25451
#1	-.00028	22.689	.03110	.00526	.00208	-.00358	.00616	.97332
#2	.00048	23.539	.03160	.01013	.00311	-.00330	.00916	.96949
#3	.00022	23.422	.03007	.00303	.00381	.00355	.00387	.96870
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606128901 Acquired: 6/25/2016 15:44:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00051	.16522	-.00360	-.00090	.00125	.00115	.00088
Stddev	.00057	.00251	.00574	.00300	.00085	.00009	.00100
%RSD	111.15	1.5196	159.60	334.58	68.243	8.1341	114.01
#1	-.00005	.16233	-.01006	-.00433	.00031	.00125	.00060
#2	.00108	.16651	.00089	.00117	.00197	.00107	.00199
#3	.00050	.16682	-.00162	.00047	.00148	.00112	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9940.0	60088.	3038.6
Stddev	4.2	150.	44.1
%RSD	.04218	.24912	1.4500
#1	9935.3	60256.	3084.8
#2	9941.7	60040.	2997.0
#3	9943.1	59969.	3034.2

Approved: June 27, 2016

Sample Name: L1606112901 Acquired: 6/25/2016 15:48:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00126	-.00046	-.00005	-.00065	.00318	.00006	4.1415	-.00016
Stddev	.00197	.00573	.00201	.00160	.00151	.00005	.0124	.00023
%RSD	156.91	1238.2	4274.9	245.00	47.612	81.871	.30015	142.20
#1	.00341	-.00015	-.00215	-.00141	.00362	.00006	4.1534	-.00032
#2	-.00045	.00511	.00185	.00118	.00150	.00012	4.1425	.00010
#3	.00081	-.00635	.00016	-.00173	.00443	.00001	4.1286	-.00026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00001	.00058	.00066	.05376	.10451	-.00223	.92501	.00486
Stddev	.00022	.00057	.00151	.01103	.06065	.00348	.08081	.00144
%RSD	3258.5	98.481	227.65	20.514	58.030	156.17	8.7365	29.653
#1	-.00025	.00121	.00094	.05934	.12513	.00172	.91232	.00510
#2	.00010	.00039	-.00097	.04105	.03624	-.00357	1.0114	.00331
#3	.00016	.00012	.00201	.06088	.15216	-.00484	.85130	.00616
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00027	3.0376	-.00008	.00578	.00044	-.00230	.00144	.21434
Stddev	.00025	.0744	.00053	.00529	.00227	.00178	.00420	.00577
%RSD	93.741	2.4506	639.89	91.432	509.94	77.275	291.44	2.6916
#1	-.00011	2.9794	.00051	.00238	.00289	-.00166	.00573	.20865
#2	-.00014	3.0119	-.00050	.00310	-.00159	-.00431	-.00267	.21418
#3	-.00056	3.1215	-.00026	.01187	.00003	-.00093	.00126	.22019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606112901 Acquired: 6/25/2016 15:48:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00020**	**.01265**	**.00270**	**.00064**	**.00099**	**.00067**	**.00092**
Stddev	.00055	.00035	.00371	.00238	.00052	.00018	.00051
%RSD	272.21	2.7272	137.50	373.62	52.293	26.328	55.163
#1	-.00083	.01259	.00376	-.00183	.00159	.00059	.00050
#2	.00008	.01234	-.00143	.00082	.00070	.00087	.00077
#3	.00015	.01303	.00576	.00292	.00069	.00054	.00148
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9969.3**	**60775.**	**3044.6**
Stddev	21.9	129.	15.2
%RSD	.22015	.21273	.49841
#1	9984.3	60920.	3056.8
#2	9944.1	60732.	3049.4
#3	9979.5	60672.	3027.6

Approved: June 27, 2016

Sample Name: L1606112902 Acquired: 6/25/2016 15:52:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00026	.00804	.00035	-.00128	.00288	.00010	2.9594	-.00006
Stddev	.00058	.00641	.00226	.00137	.00187	.00003	.0055	.00010
%RSD	221.07	79.747	653.27	107.46	64.882	33.893	.18419	164.92
#1	.00004	.00818	.00185	.00031	.00398	.00007	2.9654	.00000
#2	.00011	.00156	.00144	-.00200	.00072	.00009	2.9582	-.00001
#3	-.00094	.01439	-.00226	-.00213	.00394	.00014	2.9547	-.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00001	.00045	.00076	.08667	.10793	.00221	.95476	.02200
Stddev	.00021	.00074	.00096	.00189	.05164	.00058	.08086	.00168
%RSD	3252.3	165.35	125.16	2.1842	47.845	26.260	8.4687	7.6363
#1	.00009	.00055	.00166	.08573	.06794	.00282	.86999	.02017
#2	-.00024	-.00034	.00088	.08885	.16623	.00215	.96327	.02346
#3	.00016	.00114	-.00024	.08543	.08961	.00166	1.0310	.02238
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00025	1.4239	.00014	.00542	.00025	-.00059	.00178	.20161
Stddev	.00013	.0489	.00086	.00331	.00120	.00643	.00247	.00182
%RSD	52.086	3.4352	596.95	60.989	470.84	1086.5	138.88	.90303
#1	-.00010	1.4077	-.00013	.00438	-.00113	.00173	.00080	.20089
#2	-.00031	1.4789	-.00055	.00913	.00091	-.00786	-.00005	.20368
#3	-.00033	1.3852	.00110	.00276	.00098	.00435	.00459	.20026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606112902 Acquired: 6/25/2016 15:52:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00083	.01066	.00349	-.00281	.00049	.00074	-.00046
Stddev	.00025	.00027	.00348	.00166	.00046	.00011	.00023
%RSD	30.685	2.4916	99.744	59.148	94.747	14.382	50.877
#1	.00087	.01070	.00643	-.00461	.00075	.00069	-.00046
#2	.00056	.01090	.00439	-.00132	-.00005	.00087	-.00023
#3	.00106	.01038	-.00035	-.00252	.00076	.00067	-.00070
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9989.3	60825.	3066.0
Stddev	18.4	89.	4.0
%RSD	.18402	.14693	.13203
#1	10008.	60870.	3065.9
#2	9971.8	60722.	3062.0
#3	9987.6	60883.	3070.1

Approved: June 27, 2016

Sample Name: L1606112903 Acquired: 6/25/2016 15:56:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00034**	**-.00746**	**.00009**	**.00007**	**.00214**	**.00016**	**4.2282**	**-.00031**
Stddev	.00178	.00543	.00185	.00098	.00088	.00001	.1021	.00011
%RSD	518.14	72.757	1962.2	1427.6	41.098	3.6332	2.4147	36.326
#1	-.00212	-.00132	-.00121	.00003	.00116	.00017	4.2310	-.00019
#2	.00143	-.00942	.00221	-.00089	.00287	.00016	4.3288	-.00035
#3	-.00034	-.01164	-.00072	.00107	.00238	.00015	4.1247	-.00041
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00027**	**-.00057**	**-.00029**	**.07902**	**-.03881**	**.00121**	**1.0358**	**.00397**
Stddev	.00063	.00094	.00019	.01076	.01948	.00237	.1206	.00030
%RSD	236.86	166.12	66.025	13.622	50.186	195.96	11.648	7.4974
#1	-.00020	-.00113	-.00052	.06755	-.06059	-.00026	.93657	.00379
#2	.00099	.00052	-.00021	.08890	-.02305	-.00005	1.1701	.00431
#3	.00002	-.00109	-.00016	.08063	-.03281	.00395	1.0007	.00380
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00055**	**2.1503**	**-.00065**	**.00144**	**.00315**	**-.00171**	**-.00059**	**.27406**
Stddev	.00026	.0349	.00083	.00721	.00266	.00193	.00290	.00085
%RSD	47.062	1.6229	127.58	500.84	84.421	112.70	495.02	.31027
#1	-.00082	2.1700	-.00132	-.00288	.00271	.00020	-.00182	.27322
#2	-.00053	2.1710	.00028	-.00257	.00074	-.00366	.00273	.27492
#3	-.00030	2.1101	-.00091	.00976	.00599	-.00168	-.00267	.27403
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606112903 Acquired: 6/25/2016 15:56:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	.01267	.00761	-.00035	.00044	.00047	.00034
Stddev	.00036	.00010	.00488	.00192	.00076	.00017	.00099
%RSD	75.051	.78508	64.089	556.65	172.53	36.380	290.00
#1	.00045	.01256	.01172	-.00046	-.00009	.00039	-.00046
#2	.00086	.01267	.00891	.00163	.00010	.00066	.00145
#3	.00013	.01276	.00222	-.00221	.00131	.00034	.00003
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9986.0	60665.	3026.9
Stddev	20.7	145.	17.8
%RSD	.20708	.23844	.58822
#1	10007.	60770.	3014.6
#2	9984.8	60726.	3018.9
#3	9966.0	60500.	3047.3

Approved: June 27, 2016

Sample Name: L1606112904 Acquired: 6/25/2016 16:00:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00143**	**.00782**	**.00014**	**-.00105**	**.00318**	**.00018**	**2.9065**	**-.00009**
Stddev	.00049	.00478	.00214	.00233	.00102	.00003	.0438	.00019
%RSD	34.335	61.153	1520.7	221.14	32.071	15.659	1.5085	207.21
#1	.00199	.01026	-.00194	.00103	.00202	.00019	2.8571	-.00015
#2	.00122	.00231	.00002	-.00063	.00358	.00019	2.9407	.00012
#3	.00108	.01088	.00233	-.00357	.00394	.00014	2.9218	-.00024
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00050**	**.00045**	**.00112**	**.05226**	**.04902**	**.00077**	**.91311**	**.02436**
Stddev	.00068	.00160	.00112	.00414	.07439	.00531	.01971	.00311
%RSD	135.32	357.44	99.972	7.9272	151.77	691.64	2.1590	12.769
#1	.00007	.00219	.00019	.05089	.09310	-.00511	.93142	.02095
#2	.00129	.00012	.00080	.04898	.09083	.00221	.91566	.02510
#3	.00015	-.00096	.00236	.05692	-.03687	.00521	.89224	.02704
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00023**	**1.3676**	**-.00081**	**.00778**	**.00502**	**-.00289**	**.00478**	**.20680**
Stddev	.00041	.0202	.00063	.00381	.00153	.00221	.00206	.00531
%RSD	176.12	1.4731	77.366	49.002	30.478	76.529	43.075	2.5661
#1	-.00056	1.3492	-.00140	.01136	.00503	-.00266	.00633	.20082
#2	-.00037	1.3644	-.00015	.00819	.00655	-.00520	.00556	.20859
#3	.00023	1.3891	-.00088	.00377	.00348	-.00080	.00244	.21097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606112904 Acquired: 6/25/2016 16:00:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00019	.01047	.00111	-.00019	.00078	.00080	-.00006
Stddev	.00022	.00018	.00665	.00270	.00041	.00022	.00073
%RSD	116.67	1.7478	597.28	1390.3	52.913	27.125	1123.9
#1	.00044	.01026	.00785	.00028	.00121	.00056	.00012
#2	.00008	.01056	.00093	.00224	.00075	.00085	.00055
#3	.00005	.01059	-.00544	-.00310	.00038	.00099	-.00086
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9957.7	60557.	3030.1
Stddev	15.8	177.	8.6
%RSD	.15901	.29213	.28426
#1	9975.9	60586.	3020.4
#2	9949.3	60718.	3033.0
#3	9947.7	60368.	3036.9

Approved: June 27, 2016

Sample Name: L1606127201 Acquired: 6/25/2016 16:04:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00028	-.00887	-.00183	.00197	.00056	.00010	.95406	-.00003
Stddev	.00091	.01117	.00052	.00188	.00043	.00011	.05017	.00010
%RSD	325.46	125.87	28.528	95.488	77.565	107.58	5.2586	289.02
#1	.00022	-.02176	-.00218	.00027	.00085	.00010	.90480	.00002
#2	.00122	-.00213	-.00123	.00398	.00076	-.00001	.95228	.00003
#3	-.00060	-.00272	-.00207	.00165	.00006	.00021	1.0051	-.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00003	-.00128	.00046	.02799	-.07233	.00197	.21481	.00019
Stddev	.00061	.00045	.00022	.01180	.06961	.00022	.12156	.00200
%RSD	1838.2	35.569	48.564	42.171	96.239	11.394	56.590	1042.3
#1	-.00013	-.00080	.00028	.03989	-.13804	.00212	.34118	.00130
#2	-.00048	-.00170	.00071	.01628	.00061	.00171	.09871	-.00212
#3	.00071	-.00133	.00040	.02780	-.07955	.00208	.20454	.00139
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00067	6.5927	-.00041	.00599	.00226	-.00187	.00139	.23314
Stddev	.00039	.0563	.00074	.00538	.00192	.00575	.00169	.00199
%RSD	58.032	.85407	180.55	89.733	84.902	308.02	122.23	.85400
#1	-.00069	6.5318	-.00010	.00974	.00276	.00008	.00216	.23308
#2	-.00105	6.6428	-.00125	-.00017	.00014	-.00834	-.00056	.23517
#3	-.00027	6.6034	.00012	.00842	.00389	.00266	.00255	.23118
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127201 Acquired: 6/25/2016 16:04:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00023**	**.00331**	**.00248**	**.00096**	**.00107**	**.00085**	**.00133**
Stddev	.00044	.00056	.00254	.00297	.00078	.00012	.00063
%RSD	192.52	16.945	102.26	309.29	73.373	14.693	47.329
#1	.00017	.00395	.00360	.00246	.00196	.00094	.00062
#2	-.00070	.00291	.00428	.00288	.00074	.00089	.00180
#3	-.00016	.00306	-.00042	-.00246	.00050	.00071	.00159
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9972.6**	**60217.**	**3037.7**
Stddev	14.0	83.	14.1
%RSD	.14049	.13799	.46569
#1	9988.5	60309.	3026.4
#2	9967.3	60146.	3033.1
#3	9962.1	60196.	3053.5

Approved: June 27, 2016

Sample Name: L1606127202 Acquired: 6/25/2016 16:08:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00026**	**.01060**	**.00084**	**.01591**	**.00482**	**.00013**	**4.8874**	**.00012**
Stddev	.00147	.00116	.00295	.00196	.00045	.00002	.0512	.00031
%RSD	570.05	10.898	350.53	12.303	9.3008	12.581	1.0477	272.03
#1	-.00187	.01061	.00261	.01812	.00447	.00014	4.8299	.00013
#2	.00100	.00945	-.00256	.01523	.00533	.00015	4.9279	.00042
#3	.00009	.01176	.00248	.01439	.00467	.00011	4.9046	-.00021
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00019**	**-.00046**	**.00025**	**.03059**	**.03395**	**.00208**	**2.7028**	**-.00061**
Stddev	.00015	.00087	.00083	.01638	.06594	.00230	.1335	.00149
%RSD	81.112	189.49	337.36	53.558	194.25	110.69	4.9393	243.32
#1	.00003	-.00146	.00109	.04846	-.00968	-.00038	2.8372	-.00173
#2	.00034	.00002	-.00057	.01628	.00172	.00418	2.5702	.00108
#3	.00021	.00006	.00022	.02703	.10981	.00243	2.7010	-.00120
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00026**	**13.742**	**-.00085**	**.00227**	**.00279**	**-.00269**	**-.00193**	**.47645**
Stddev	.00056	.155	.00030	.00745	.00217	.00177	.00170	.00665
%RSD	217.74	1.1297	35.156	328.49	77.588	65.958	88.159	1.3954
#1	-.00002	13.587	-.00065	-.00435	.00526	-.00088	-.00382	.46927
#2	-.00089	13.898	-.00120	.00082	.00118	-.00276	-.00050	.47768
#3	.00014	13.742	-.00071	.01033	.00194	-.00443	-.00148	.48240
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127202 Acquired: 6/25/2016 16:08:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00018**	**.05092**	**.00177**	**.00024**	**.00137**	**.00071**	**.00162**
Stddev	.00083	.00100	.00998	.00193	.00052	.00009	.00084
%RSD	474.40	1.9658	565.04	789.52	37.663	12.612	51.686
#1	-.00049	.05003	.00781	-.00019	.00196	.00074	.00252
#2	-.00081	.05200	-.00975	.00235	.00120	.00078	.00148
#3	.00077	.05072	.00724	-.00143	.00097	.00061	.00086
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9920.3**	**59957.**	**3053.1**
Stddev	9.3	97.	28.1
%RSD	.09367	.16197	.92019
#1	9910.9	59874.	3056.5
#2	9929.5	60064.	3023.5
#3	9920.6	59932.	3079.4

Approved: June 27, 2016

Sample Name: L1606127901 Acquired: 6/25/2016 16:12:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00079**	**.23343**	**-.00035**	**.00697**	**.01048**	**.00010**	**5.5956**	**-.00004**
Stddev	.00151	.00808	.00093	.00046	.00038	.00003	.0897	.00007
%RSD	191.07	3.4604	268.33	6.6127	3.6255	27.106	1.6038	155.88
#1	-.00012	.24183	.00073	.00738	.01041	.00014	5.5127	-.00007
#2	.00027	.23274	-.00087	.00647	.01090	.00009	5.5834	-.00010
#3	-.00251	.22572	-.00090	.00704	.01015	.00009	5.6909	.00003
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00030**	**.00040**	**.00104**	**.37949**	**.02999**	**.00120**	**2.2214**	**.04375**
Stddev	.00022	.00063	.00117	.02000	.03448	.00297	.1047	.00042
%RSD	72.888	155.32	112.12	5.2704	114.98	248.44	4.7142	.95613
#1	.00020	.00102	-.00030	.35646	.01685	.00172	2.1668	.04335
#2	.00015	-.00023	.00181	.39248	.06911	.00387	2.1552	.04372
#3	.00056	.00042	.00162	.38953	.00401	-.00200	2.3421	.04419
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**.40372**	**.00002**	**.01800**	**.00182**	**-.00229**	**.00534**	**.76285**
Stddev	.00016	.02832	.00029	.00724	.00294	.00133	.00501	.02556
%RSD	124.97	7.0152	1295.9	40.230	161.43	58.036	93.812	3.3508
#1	-.00031	.38955	.00035	.01067	.00373	-.00090	.01054	.73582
#2	-.00001	.38528	-.00008	.01820	-.00157	-.00355	.00492	.76609
#3	-.00007	.43633	-.00020	.02515	.00330	-.00243	.00055	.78664
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127901 Acquired: 6/25/2016 16:12:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00000**	**.02802**	**.00639**	**-.00424**	**.00104**	**.00686**	**-.00053**
Stddev	.00059	.00084	.00088	.00033	.00039	.00027	.00162
%RSD	12664.	3.0023	13.821	7.8780	37.381	4.0081	307.76
#1	.00044	.02706	.00703	-.00390	.00138	.00661	.00100
#2	.00022	.02835	.00538	-.00425	.00062	.00681	-.00222
#3	-.00068	.02864	.00675	-.00456	.00114	.00715	-.00036
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9995.1**	**60569.**	**3072.5**
Stddev	18.8	83.	17.9
%RSD	.18798	.13702	.58354
#1	10017.	60498.	3093.1
#2	9983.0	60660.	3060.2
#3	9985.6	60548.	3064.2

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 16:16:03 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41279**	**10.433**	**.40298**	**.50323**	**1.0184**	**.05236**	**10.146**
Stddev	.00062	.023	.00491	.00242	.0037	.00006	.030
%RSD	.14995	.22399	1.2186	.48043	.36729	.12018	.29415
#1	.41231	10.415	.40384	.50055	1.0220	.05229	10.131
#2	.41349	10.459	.39769	.50390	1.0146	.05241	10.126
#3	.41257	10.425	.40740	.50524	1.0186	.05238	10.180
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05206**	**.20294**	**.51645**	**.52099**	**4.0952**	**49.051**	**1.0295**
Stddev	.00043	.00065	.00457	.00121	.0025	.142	.0088
%RSD	.83054	.31937	.88415	.23159	.06120	.28862	.85476
#1	.05248	.20360	.51186	.52157	4.0943	49.074	1.0305
#2	.05210	.20291	.52099	.52180	4.0933	48.900	1.0203
#3	.05161	.20230	.51649	.51960	4.0980	49.180	1.0378
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.7277**	**.51414**	**1.0402**	**49.935**	**.51586**	**10.134**	**.51189**
Stddev	.0462	.00321	.0047	.095	.00165	.031	.00166
%RSD	.47447	.62359	.45388	.19063	.31918	.30495	.32486
#1	9.7465	.51782	1.0451	50.021	.51766	10.168	.51381
#2	9.6751	.51192	1.0398	49.833	.51549	10.123	.51086
#3	9.7615	.51270	1.0357	49.951	.51442	10.109	.51101
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 16:16:03 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2153**	**.40661**	**5.2188**	**1.0211**	**1.0187**	**1.0271**	**.52415**
Stddev	.0058	.00386	.0120	.0013	.0031	.0069	.00046
%RSD	.47598	.95000	.22957	.12690	.30817	.67428	.08834
#1	1.2162	.40345	5.2255	1.0218	1.0207	1.0286	.52363
#2	1.2206	.41091	5.2258	1.0219	1.0150	1.0195	.52433
#3	1.2091	.40546	5.2049	1.0196	1.0203	1.0331	.52450
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0611**	**1.0269**	F **2.9985**
Stddev	.0021	.0010	.0124
%RSD	.19830	.09638	.41405
#1	1.0631	1.0270	3.0047
#2	1.0614	1.0279	2.9842
#3	1.0589	1.0259	3.0065
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9559.6**	**57599.**	**2986.8**
Stddev	4.9	130.	10.5
%RSD	.05148	.22635	.35191
#1	9564.0	57612.	2977.9
#2	9560.6	57462.	2998.4
#3	9554.3	57722.	2984.0

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 16:19:40 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00185**	**.00626**	**-.00023**	**.00044**	**.00278**	**.00005**	**-.04464**	**.00001**
Stddev	.00025	.00998	.00075	.00197	.00045	.00006	.01589	.00008
%RSD	13.751	159.28	328.05	451.88	16.156	122.09	35.589	834.92
#1	.00158	.01229	-.00019	.00230	.00236	.00010	-.04122	-.00004
#2	.00190	-.00525	.00050	.00063	.00326	-.00002	-.03074	-.00004
#3	.00208	.01176	-.00099	-.00162	.00274	.00006	-.06195	.00010
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00039**	**.00030**	**.00108**	**.01891**	**-.08883**	**.00028**	**.04170**	**-.00007**
Stddev	.00014	.00149	.00123	.02463	.04133	.00158	.04172	.00069
%RSD	37.231	490.44	114.38	130.22	46.533	569.52	100.03	937.33
#1	-.00023	.00135	.00225	.04102	-.10527	-.00061	.02819	.00053
#2	-.00043	.00096	-.00021	.02335	-.11941	.00210	.00842	.00008
#3	-.00050	-.00140	.00119	-.00763	-.04180	-.00066	.08850	-.00082
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00343**	**.00251**	**-.00012**	**.00363**	**.00401**	**.00489**	**.00879**	**-.00487**
Stddev	.00038	.01645	.00050	.00608	.00167	.00193	.00298	.00028
%RSD	11.063	654.49	431.61	167.28	41.713	39.486	33.953	5.7618
#1	.00299	.00408	-.00004	.00277	.00588	.00378	.00970	-.00459
#2	.00370	-.01467	-.00065	-.00196	.00352	.00712	.01122	-.00515
#3	.00359	.01813	.00034	.01010	.00264	.00378	.00546	-.00487
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 16:19:40 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00028	.00057	.00214	-.00157	.00060	-.00006	.00421
Stddev	.00055	.00026	.00424	.00416	.00034	.00004	.00038
%RSD	193.42	45.468	198.11	264.68	55.851	78.634	9.1291
#1	-.00056	.00032	.00612	.00235	.00061	-.00008	.00456
#2	-.00063	.00083	-.00232	-.00593	.00093	-.00001	.00426
#3	.00035	.00056	.00262	-.00114	.00026	-.00008	.00380
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9751.1	59011.	2998.0
Stddev	16.0	92.	16.9
%RSD	.16409	.15669	.56327
#1	9735.8	59057.	2997.7
#2	9767.8	59072.	2981.2
#3	9749.7	58905.	3015.0

Approved: June 27, 2016

Sample Name: L1606127905 Acquired: 6/25/2016 16:23:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00112**	**.42892**	**.00254**	**.00885**	**.00931**	**.00014**	**4.1101**	**-.00014**
Stddev	.00085	.00960	.00110	.00026	.00054	.00008	.0226	.00009
%RSD	76.091	2.2373	43.205	2.9736	5.7541	61.416	.54879	60.631
#1	.00187	.41894	.00336	.00888	.00991	.00015	4.0848	-.00017
#2	.00019	.43809	.00129	.00858	.00890	.00022	4.1178	-.00005
#3	.00131	.42973	.00298	.00910	.00911	.00005	4.1279	-.00021
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00064**	**.00072**	**.00093**	**.46806**	**-.04215**	**.00375**	**2.3175**	**.01534**
Stddev	.00025	.00122	.00082	.01630	.03381	.00237	.0180	.00171
%RSD	38.393	169.05	88.133	3.4825	80.210	63.236	.77530	11.156
#1	.00092	.00051	.00011	.45372	-.08111	.00533	2.3322	.01684
#2	.00057	-.00038	.00092	.46467	-.02046	.00102	2.2974	.01572
#3	.00044	.00203	.00175	.48579	-.02489	.00491	2.3228	.01348
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00042**	**.79830**	**.00036**	**.03423**	**.00164**	**-.00327**	**.00568**	**1.0453**
Stddev	.00055	.00874	.00054	.00334	.00066	.00436	.00399	.0205
%RSD	130.46	1.0949	148.73	9.7489	39.962	133.66	70.190	1.9601
#1	-.00012	.80324	-.00001	.03061	.00234	-.00815	.00802	1.0252
#2	.00040	.78821	.00099	.03719	.00155	-.00191	.00794	1.0445
#3	.00099	.80344	.00012	.03489	.00104	.00026	.00108	1.0662
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127905 Acquired: 6/25/2016 16:23:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.01799	.00482	-.00230	.00189	.01109	.00007
Stddev	.00065	.00037	.00226	.00232	.00074	.00029	.00211
%RSD	715.49	2.0334	47.005	100.92	39.126	2.6113	3180.6
#1	-.00003	.01830	.00310	.00003	.00148	.01082	.00082
#2	.00080	.01809	.00396	-.00232	.00274	.01107	-.00231
#3	-.00049	.01759	.00738	-.00462	.00144	.01140	.00169
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9972.5	60381.	3064.6
Stddev	33.3	69.	2.8
%RSD	.33435	.11363	.09242
#1	10010.	60351.	3061.5
#2	9961.5	60333.	3067.1
#3	9946.1	60460.	3065.2

Approved: June 27, 2016

Sample Name: L1606114801 Acquired: 6/25/2016 16:27:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.03462	.00065	.00821	.00587	.00014	1.4096	-.00013
Stddev	.00034	.00628	.00229	.00098	.00178	.00003	.0346	.00033
%RSD	57.407	18.141	351.58	11.960	30.235	19.203	2.4561	248.47
#1	.00054	.03990	-.00071	.00924	.00790	.00011	1.3926	.00011
#2	.00095	.02767	-.00063	.00729	.00463	.00014	1.4495	-.00000
#3	.00028	.03630	.00329	.00809	.00508	.00016	1.3868	-.00051
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.00001	.00030	.04215	.21418	.00024	.69121	.00161
Stddev	.00012	.00038	.00134	.02269	.07335	.00510	.07378	.00189
%RSD	127.75	2953.4	439.05	53.849	34.250	2116.0	10.674	117.34
#1	.00006	.00044	.00117	.03668	.16871	.00256	.67687	-.00053
#2	-.00001	-.00026	.00097	.06707	.29880	.00377	.77111	.00305
#3	.00022	-.00014	-.00123	.02268	.17502	-.00561	.62565	.00230
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00000	3.4091	-.00039	.01019	-.00019	-.00195	.00226	.25600
Stddev	.00027	.0299	.00017	.00142	.00263	.00116	.00241	.00788
%RSD	8569.8	.87827	42.602	13.923	1405.5	59.228	106.91	3.0774
#1	.00029	3.3757	-.00048	.01168	-.00018	-.00094	.00166	.24819
#2	-.00005	3.4183	-.00020	.01005	.00244	-.00171	.00492	.25587
#3	-.00025	3.4335	-.00049	.00885	-.00282	-.00321	.00020	.26394
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606114801 Acquired: 6/25/2016 16:27:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00043**	**.01395**	**.00070**	**-.00087**	**.00132**	**.00137**	**.00142**
Stddev	.00051	.00044	.00486	.00154	.00026	.00005	.00170
%RSD	118.61	3.1779	695.33	177.20	19.933	3.3203	119.70
#1	-.00063	.01431	.00581	-.00155	.00155	.00135	.00002
#2	.00015	.01345	.00013	-.00194	.00103	.00135	.00093
#3	-.00082	.01408	-.00385	.00089	.00136	.00143	.00331
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9913.2**	**60132.**	**3023.4**
Stddev	14.6	83.	23.3
%RSD	.14679	.13818	.77035
#1	9929.6	60151.	3030.8
#2	9901.6	60041.	2997.3
#3	9908.5	60204.	3042.1

Approved: June 27, 2016

Sample Name: L1606114802 Acquired: 6/25/2016 16:31:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00158	-.00408	.00026	.00633	.00904	.00015	2.7703	-.00011
Stddev	.00108	.00521	.00166	.00204	.00121	.00009	.0227	.00014
%RSD	68.438	127.59	628.67	32.297	13.379	59.313	.82092	123.10
#1	.00072	-.00923	-.00155	.00429	.00969	.00008	2.7476	-.00002
#2	.00123	.00119	.00169	.00631	.00978	.00025	2.7930	-.00027
#3	.00279	-.00421	.00065	.00838	.00764	.00012	2.7703	-.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00034	.00002	.00060	.02383	.06717	.00040	2.3757	.00083
Stddev	.00028	.00081	.00057	.01240	.09428	.00275	.0633	.00197
%RSD	81.635	5221.9	93.832	52.028	140.36	693.38	2.6650	237.49
#1	.00044	-.00078	.00046	.02298	.16159	-.00259	2.3743	.00224
#2	.00003	-.00000	.00012	.01188	-.02697	.00095	2.4397	.00167
#3	.00055	.00084	.00122	.03663	.06688	.00283	2.3131	-.00142
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00011	6.0259	-.00040	.00438	.00227	-.00140	.00185	.32995
Stddev	.00024	.0927	.00045	.00883	.00074	.00370	.00550	.00769
%RSD	208.31	1.5391	111.52	201.54	32.728	264.56	298.09	2.3316
#1	.00039	5.9438	-.00070	.01456	.00143	-.00567	.00483	.32126
#2	-.00003	6.1265	.00011	-.00011	.00282	.00058	.00521	.33270
#3	-.00002	6.0074	-.00061	-.00130	.00258	.00089	-.00450	.33589
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606114802 Acquired: 6/25/2016 16:31:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00030	.01996	.00356	-.00001	.00116	.00095	-.00024
Stddev	.00040	.00029	.00718	.00102	.00055	.00026	.00029
%RSD	136.64	1.4402	201.93	7537.5	47.304	27.165	121.82
#1	-.00006	.02008	.00861	-.00117	.00177	.00075	-.00049
#2	.00022	.02016	-.00466	.00037	.00099	.00086	-.00028
#3	.00074	.01963	.00672	.00076	.00071	.00124	.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9925.6	60099.	3031.8
Stddev	7.7	64.	18.1
%RSD	.07799	.10696	.59652
#1	9922.2	60109.	3022.9
#2	9920.2	60157.	3019.9
#3	9934.5	60030.	3052.6

Approved: June 27, 2016

Sample Name: L1606115001 Acquired: 6/25/2016 16:35:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00166	.05641	.00050	.00663	.00676	.00016	1.7818	.00015
Stddev	.00049	.00787	.00304	.00311	.00055	.00011	.0479	.00031
%RSD	29.631	13.947	601.42	46.927	8.1582	71.613	2.6886	203.48
#1	.00200	.06510	-.00294	.00991	.00739	.00005	1.7526	.00038
#2	.00109	.05435	.00166	.00372	.00635	.00014	1.8371	-.00020
#3	.00188	.04978	.00279	.00626	.00655	.00027	1.7557	.00027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00007	.00048	.00100	.04970	.00137	.00112	1.1341	.00143
Stddev	.00049	.00051	.00115	.01548	.09645	.00218	.0897	.00096
%RSD	745.24	106.17	114.47	31.142	7038.5	195.00	7.9106	67.253
#1	-.00002	.00105	.00223	.06756	-.03266	.00315	1.0361	.00043
#2	-.00057	.00034	.00081	.04145	.11022	-.00118	1.2123	.00235
#3	.00040	.00005	-.00004	.04010	-.07346	.00137	1.1537	.00152
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00027	5.9717	-.00015	.00950	-.00176	-.00302	.00622	.28058
Stddev	.00014	.0718	.00043	.00093	.00081	.00221	.00277	.00537
%RSD	53.804	1.2031	276.46	9.7789	45.909	73.167	44.558	1.9140
#1	-.00010	5.9074	-.00014	.01003	-.00249	-.00320	.00355	.27441
#2	-.00036	6.0493	-.00058	.01005	-.00089	-.00072	.00603	.28316
#3	-.00034	5.9584	.00027	.00843	-.00189	-.00513	.00908	.28418
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115001 Acquired: 6/25/2016 16:35:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00008	.01287	.00136	-.00194	.00110	.00114	.00062
Stddev	.00058	.00017	.00295	.00426	.00072	.00013	.00087
%RSD	762.02	1.3177	217.51	219.83	65.867	11.722	140.33
#1	.00015	.01301	.00162	-.00450	.00182	.00119	.00034
#2	-.00054	.01292	.00417	-.00429	.00110	.00099	-.00008
#3	.00061	.01268	-.00172	.00298	.00038	.00124	.00159
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9919.3	60059.	3016.8
Stddev	34.5	56.	13.7
%RSD	.34767	.09285	.45386
#1	9942.2	60062.	3013.8
#2	9936.0	60114.	3004.8
#3	9879.6	60002.	3031.7

Approved: June 27, 2016

Sample Name: L1606115102 Acquired: 6/25/2016 16:39:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00105	.00214	.00018	.00362	.00526	.00013	1.7820	.00006
Stddev	.00166	.00346	.00084	.00176	.00021	.00005	.0144	.00011
%RSD	158.53	161.68	476.67	48.690	4.0116	41.814	.80931	182.53
#1	.00213	.00451	-.00003	.00223	.00549	.00009	1.7780	.00003
#2	-.00086	-.00183	-.00054	.00560	.00521	.00010	1.7700	-.00003
#3	.00188	.00375	.00110	.00302	.00507	.00019	1.7980	.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00024	.00030	.00081	.02568	-.04054	.00233	1.0068	.00278
Stddev	.00009	.00056	.00053	.02210	.04687	.00374	.0618	.00296
%RSD	39.196	184.13	65.165	86.089	115.61	160.62	6.1428	106.56
#1	.00032	.00005	.00046	.05079	-.00224	.00526	1.0696	.00587
#2	.00013	-.00008	.00055	.00916	-.02657	.00361	.94591	-.00003
#3	.00026	.00094	.00141	.01708	-.09281	-.00188	1.0048	.00250
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00000	4.8272	.00042	.00764	.00031	-.00602	.00252	.26699
Stddev	.00027	.0498	.00058	.00427	.00107	.00140	.00418	.00420
%RSD	7088.8	1.0326	136.65	55.913	346.71	23.202	166.24	1.5735
#1	-.00018	4.7713	.00031	.01044	.00076	-.00742	-.00226	.26233
#2	-.00015	4.8672	-.00009	.00272	.00108	-.00462	.00551	.26815
#3	.00031	4.8430	.00105	.00974	-.00091	-.00602	.00431	.27049
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115102 Acquired: 6/25/2016 16:39:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00036	.01171	-.00105	-.00167	.00048	.00111	-.00010
Stddev	.00036	.00032	.00211	.00218	.00034	.00004	.00053
%RSD	98.761	2.7498	200.92	130.51	71.305	3.3481	503.34
#1	-.00064	.01138	-.00295	-.00411	.00028	.00107	-.00046
#2	-.00048	.01174	.00122	-.00099	.00029	.00113	.00050
#3	.00004	.01202	-.00143	.00009	.00088	.00113	-.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9924.7	60209.	3035.9
Stddev	21.0	164.	13.4
%RSD	.21110	.27319	.44278
#1	9945.2	60212.	3050.8
#2	9925.5	60043.	3024.7
#3	9903.3	60372.	3032.1

Approved: June 27, 2016

Sample Name: L1606115105 Acquired: 6/25/2016 16:43:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00211	.00219	.00055	.00586	.00711	.00018	2.2105	.00001
Stddev	.00101	.00795	.00092	.00071	.00025	.00010	.0174	.00025
%RSD	47.839	363.29	167.06	12.119	3.4464	53.942	.78551	4197.7
#1	.00096	-.00697	.00162	.00634	.00685	.00013	2.1935	-.00020
#2	.00251	.00621	-.00001	.00619	.00715	.00012	2.2282	.00029
#3	.00286	.00732	.00005	.00504	.00733	.00030	2.2098	-.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00003	-.00001	.00002	.02737	.01189	.00109	1.3118	.00125
Stddev	.00029	.00051	.00146	.02301	.05921	.00366	.0639	.00239
%RSD	827.06	7244.7	6104.0	84.080	497.84	335.35	4.8727	191.19
#1	.00035	.00040	.00134	.00788	.02968	-.00236	1.3142	.00227
#2	-.00003	.00015	.00027	.02147	.06017	.00070	1.3745	.00297
#3	-.00022	-.00058	-.00154	.05276	-.05417	.00493	1.2467	-.00148
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00023	6.1621	.00034	.00603	.00215	-.00121	.00114	.31741
Stddev	.00059	.0852	.00146	.00590	.00148	.00098	.00444	.00674
%RSD	252.32	1.3824	433.95	97.933	68.608	80.696	390.40	2.1227
#1	-.00086	6.0786	-.00133	.00455	.00384	-.00161	.00564	.31056
#2	-.00015	6.2489	.00098	.00101	.00153	-.00010	.00100	.31765
#3	.00031	6.1588	.00136	.01253	.00108	-.00192	-.00323	.32403
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115105 Acquired: 6/25/2016 16:43:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00016**	**.01487**	**-.00053**	**-.00166**	**.00081**	**.00090**	**.00078**
Stddev	.00083	.00056	.00515	.00250	.00073	.00004	.00134
%RSD	533.32	3.7472	973.66	150.50	89.693	4.1812	172.14
#1	-.00110	.01433	.00470	-.00448	.00103	.00087	.00219
#2	.00015	.01545	-.00069	-.00074	.00000	.00094	.00064
#3	.00048	.01482	-.00560	.00025	.00141	.00089	-.00049
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9886.5**	**59788.**	**3007.6**
Stddev	13.0	185.	33.5
%RSD	.13157	.30931	1.1137
#1	9901.4	60001.	3037.0
#2	9880.8	59677.	2971.1
#3	9877.4	59685.	3014.8

Approved: June 27, 2016

Sample Name: L1606115401 Acquired: 6/25/2016 16:47:06 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573707-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00038	.00056	-.00040	.08474	.00202	.00013	25.077	-.00002
Stddev	.00048	.01146	.00100	.00137	.00093	.00006	.183	.00004
%RSD	125.93	2038.5	249.32	1.6179	46.237	47.434	.73077	179.25
#1	-.00014	.00766	-.00156	.08632	.00308	.00006	25.288	-.00005
#2	.00079	.00669	.00017	.08409	.00131	.00017	24.966	-.00003
#3	.00048	-.01266	.00019	.08382	.00167	.00016	24.976	.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00025	-.00049	.00090	.03824	.13932	.00586	5.7156	.01284
Stddev	.00012	.00049	.00033	.01141	.04539	.00340	.0899	.00260
%RSD	49.754	99.993	36.739	29.841	32.583	57.965	1.5726	20.218
#1	.00011	-.00085	.00123	.04902	.09367	.00206	5.7030	.01056
#2	.00028	-.00069	.00089	.02629	.18446	.00861	5.6327	.01230
#3	.00036	.00007	.00057	.03940	.13984	.00691	5.8112	.01567
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	48.626	-.00075	.01117	.00114	-.00491	.00349	.27480
Stddev	.00026	.111	.00066	.00745	.00133	.00248	.00272	.00312
%RSD	52.573	.22800	87.964	66.704	116.26	50.491	77.811	1.1358
#1	.00024	48.499	-.00062	.01513	.00176	-.00250	.00652	.27222
#2	.00075	48.707	-.00147	.01582	.00206	-.00745	.00269	.27392
#3	.00048	48.670	-.00016	.00258	-.00038	-.00478	.00127	.27827
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115401 Acquired: 6/25/2016 16:47:06 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573707-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.45176	-.00366	.00143	.00149	.00078	.00089
Stddev	.00028	.00268	.00362	.00047	.00058	.00004	.00106
%RSD	304.19	.59263	98.732	33.094	39.124	5.2122	119.26
#1	.00040	.44867	-.00000	.00088	.00209	.00083	-.00013
#2	-.00012	.45326	-.00375	.00168	.00093	.00076	.00081
#3	-.00001	.45335	-.00723	.00172	.00145	.00075	.00200
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9797.1	59243.	3012.9
Stddev	17.5	107.	17.7
%RSD	.17826	.18036	.58631
#1	9776.9	59361.	2997.1
#2	9806.9	59214.	3009.7
#3	9807.4	59153.	3032.0

Approved: June 27, 2016

Sample Name: L1606115402S Acquired: 6/25/2016 16:51:01 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: ~~WG573994-~~04 WG573707-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.01309**	**.25824**	**.01110**	**.13590**	**.02825**	**.00141**	**24.615**	**.00126**
Stddev	.00068	.00760	.00152	.00169	.00035	.00008	.204	.00013
%RSD	5.1924	2.9441	13.682	1.2425	1.2428	5.6800	.82995	10.655
#1	.01329	.24951	.01277	.13422	.02816	.00148	24.549	.00132
#2	.01365	.26341	.01072	.13589	.02864	.00133	24.844	.00111
#3	.01234	.26179	.00981	.13760	.02795	.00143	24.452	.00136
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00516**	**.01347**	**.01556**	**.12663**	**1.3534**	**.03179**	**5.7750**	**.02544**
Stddev	.00029	.00054	.00043	.02761	.0228	.00410	.0699	.00285
%RSD	5.6307	4.0382	2.7758	21.808	1.6862	12.889	1.2101	11.189
#1	.00539	.01407	.01514	.12414	1.3303	.03536	5.8428	.02774
#2	.00525	.01301	.01552	.15541	1.3541	.02732	5.7789	.02226
#3	.00483	.01334	.01600	.10034	1.3759	.03267	5.7032	.02633
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02479**	**48.072**	**.01332**	**.27819**	**.01575**	**.03066**	**.01661**	**.41681**
Stddev	.00011	.553	.00070	.01153	.00513	.00307	.00921	.00803
%RSD	.43243	1.1513	5.2238	4.1456	32.553	10.027	55.466	1.9271
#1	.02491	47.545	.01404	.27355	.01630	.02789	.02173	.40846
#2	.02475	48.649	.01266	.26969	.02058	.03397	.02212	.41747
#3	.02470	48.023	.01326	.29132	.01037	.03013	.00597	.42449
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115402S Acquired: 6/25/2016 16:51:01 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: ~~WG573994-04~~ WG573707-07

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02366**	**.45706**	**.02085**	**.01353**	**.02750**	**.02826**	**.00274**
Stddev	.00062	.00423	.00301	.00109	.00044	.00035	.00077
%RSD	2.6070	.92537	14.446	8.0746	1.5967	1.2498	28.151
#1	.02338	.45248	.01904	.01478	.02755	.02786	.00198
#2	.02436	.46082	.01917	.01306	.02704	.02846	.00352
#3	.02323	.45789	.02432	.01275	.02792	.02848	.00271
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9814.2**	**59172.**	**3008.6**
Stddev	14.0	133.	22.5
%RSD	.14277	.22528	.74717
#1	9829.8	59323.	3030.6
#2	9810.3	59073.	2985.6
#3	9802.6	59119.	3009.6

Approved: June 27, 2016

Sample Name: L1606115403SD Acquired: 6/25/2016 16:54:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: ~~WG573994-05~~ WG573707-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01239	.28748	.01145	.14150	.02764	.00148	24.958	.00130
Stddev	.00114	.00888	.00065	.00093	.00147	.00009	.077	.00038
%RSD	9.2124	3.0872	5.6507	.65421	5.3298	6.1781	.30772	29.306
#1	.01125	.28330	.01179	.14053	.02888	.00144	25.006	.00087
#2	.01353	.28146	.01070	.14159	.02802	.00141	24.999	.00161
#3	.01238	.29767	.01185	.14238	.02601	.00158	24.870	.00142
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00611	.01431	.01527	.13086	1.3857	.03280	5.7817	.02287
Stddev	.00054	.00136	.00145	.01260	.0370	.00239	.0561	.00299
%RSD	8.8509	9.4861	9.4751	9.6283	2.6697	7.2716	.97114	13.075
#1	.00671	.01585	.01436	.14435	1.3780	.03556	5.7473	.02433
#2	.00566	.01329	.01452	.11939	1.4259	.03138	5.8465	.01943
#3	.00597	.01379	.01694	.12885	1.3532	.03147	5.7513	.02484
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02737	48.881	.01504	.29126	.01658	.02795	.01955	.43672
Stddev	.00021	.600	.00152	.00741	.00249	.00243	.00943	.00693
%RSD	.75532	1.2279	10.076	2.5426	15.013	8.6768	48.242	1.5861
#1	.02720	48.195	.01331	.29382	.01763	.03060	.01049	.42891
#2	.02732	49.311	.01567	.29705	.01373	.02742	.02931	.43914
#3	.02760	49.136	.01613	.28292	.01836	.02584	.01885	.44211
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115403SD Acquired: 6/25/2016 16:54:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: ~~WG573994-05~~ WG573707-08

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02590	.47018	.02569	.01379	.02893	.02996	.00113
Stddev	.00080	.00475	.00439	.00029	.00110	.00076	.00027
%RSD	3.0868	1.0106	17.097	2.0910	3.8168	2.5433	23.745
#1	.02499	.46471	.02819	.01349	.02776	.02918	.00138
#2	.02647	.47252	.02061	.01406	.02908	.03000	.00115
#3	.02625	.47331	.02825	.01382	.02996	.03071	.00085
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9727.4	58906.	3022.7
Stddev	40.6	23.	25.7
%RSD	.41767	.03969	.84864
#1	9765.9	58911.	3039.9
#2	9731.2	58926.	2993.2
#3	9685.0	58880.	3035.0

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 16:58:48 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.41361	10.445	.40516	.50612	1.0259	.05224	10.203
Stddev	.00260	.021	.00144	.00287	.0033	.00020	.046
%RSD	.62922	.20256	.35469	.56757	.31739	.38953	.44884
#1	.41655	10.462	.40524	.50908	1.0271	.05242	10.224
#2	.41159	10.421	.40656	.50592	1.0284	.05202	10.235
#3	.41269	10.451	.40369	.50335	1.0222	.05229	10.150
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05215	.20471	.51698	.52477	4.0648	49.423	1.0320
Stddev	.00013	.00021	.00148	.00103	.0369	.189	.0006
%RSD	.23994	.10391	.28587	.19687	.90838	.38306	.06170
#1	.05201	.20470	.51640	.52461	4.1007	49.428	1.0316
#2	.05218	.20450	.51865	.52588	4.0269	49.610	1.0327
#3	.05226	.20492	.51587	.52383	4.0668	49.231	1.0316
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.8590	.51551	1.0404	50.331	.51945	10.229	.51430
Stddev	.0965	.00144	.0039	.173	.00106	.010	.00570
%RSD	.97927	.27876	.37864	.34422	.20472	.09385	1.1081
#1	9.9705	.51433	1.0396	50.428	.51822	10.232	.51032
#2	9.8060	.51711	1.0369	50.435	.52013	10.218	.51175
#3	9.8006	.51508	1.0447	50.131	.51999	10.236	.52083
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 16:58:48 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2288**	**.41352**	**5.2624**	**1.0315**	**1.0269**	**1.0279**	**.53017**
Stddev	.0104	.00168	.0015	.0026	.0016	.0023	.00514
%RSD	.84208	.40618	.02811	.25583	.15527	.22255	.96993
#1	1.2288	.41198	5.2628	1.0285	1.0283	1.0298	.53564
#2	1.2391	.41326	5.2608	1.0336	1.0273	1.0286	.52543
#3	1.2184	.41531	5.2637	1.0323	1.0252	1.0254	.52943
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0564**	**1.0308**	F **3.0180**
Stddev	.0015	.0001	.0119
%RSD	.14240	.01349	.39369
#1	1.0580	1.0307	3.0121
#2	1.0561	1.0310	3.0317
#3	1.0550	1.0308	3.0102
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9528.1**	**57486.**	**2960.8**
Stddev	9.3	201.	17.3
%RSD	.09776	.34979	.58425
#1	9517.6	57273.	2962.9
#2	9535.3	57514.	2942.6
#3	9531.5	57672.	2977.0

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 17:02:25 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00045**	**.00226**	**.00067**	**.00055**	**.00185**	**.00013**	**-.07167**	**-.00014**
Stddev	.00085	.00279	.00119	.00264	.00042	.00005	.02175	.00033
%RSD	189.88	123.24	177.00	483.17	22.573	35.869	30.346	234.73
#1	.00019	-.00053	.00152	-.00019	.00233	.00018	-.09322	-.00050
#2	-.00141	.00505	-.00069	-.00165	.00153	.00012	-.04973	.00014
#3	-.00012	.00226	.00119	.00348	.00171	.00009	-.07206	-.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00002**	**.00057**	**.00076**	**.03056**	**-.09613**	**.00228**	**-.13259**	**.00051**
Stddev	.00039	.00056	.00130	.01338	.02053	.00471	.02062	.00136
%RSD	1720.9	98.019	170.60	43.800	21.361	206.77	15.553	268.18
#1	.00020	.00017	-.00074	.01806	-.09726	-.00131	-.14855	-.00092
#2	.00020	.00121	.00148	.02892	-.07505	.00054	-.10931	.00179
#3	-.00047	.00034	.00155	.04468	-.11607	.00761	-.13993	.00066
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00325**	**.04047**	**-.00081**	**.00150**	**.00148**	**.00530**	**.00256**	**-.00673**
Stddev	.00022	.01175	.00096	.00183	.00289	.00260	.00237	.00167
%RSD	6.7467	29.033	119.00	121.89	195.62	48.929	92.639	24.783
#1	.00325	.05342	-.00016	.00278	.00470	.00570	.00309	-.00596
#2	.00303	.03049	-.00036	.00231	.00063	.00768	-.00003	-.00558
#3	.00347	.03751	-.00192	-.00059	-.00090	.00253	.00461	-.00864
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 17:02:25 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00041**	**.00044**	**.00426**	**-.00209**	**.00094**	**.00001**	**.00353**
Stddev	.00052	.00045	.00463	.00321	.00096	.00021	.00070
%RSD	128.76	102.93	108.80	153.83	102.68	1611.6	19.702
#1	.00048	.00069	.00959	-.00418	.00024	-.00012	.00320
#2	.00089	.00071	.00120	.00161	.00204	-.00009	.00306
#3	-.00015	-.00008	.00198	-.00369	.00053	.00025	.00433
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9729.2**	**58924.**	**2982.2**
Stddev	14.1	124.	20.8
%RSD	.14444	.21107	.69802
#1	9725.0	58912.	2962.2
#2	9744.9	58806.	3003.8
#3	9717.8	59054.	2980.7

Approved: June 27, 2016

Sample Name: L1606115501 Acquired: 6/25/2016 17:06:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00033	.00180	.00125	.01161	.01032	.00011	2.3136	-.00007
Stddev	.00230	.00043	.00160	.00137	.00099	.00007	.0269	.00019
%RSD	694.54	23.757	128.16	11.776	9.5715	61.393	1.1630	281.96
#1	.00225	.00218	.00194	.01272	.00925	.00003	2.3397	.00009
#2	-.00214	.00134	.00239	.01008	.01120	.00016	2.2860	-.00001
#3	-.00111	.00188	-.00058	.01203	.01051	.00015	2.3151	-.00027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00005	.00014	-.00018	.02935	.09569	-.00047	1.4383	.08572
Stddev	.00040	.00048	.00052	.00252	.08702	.00314	.0635	.00277
%RSD	831.42	341.03	289.17	8.5796	90.931	664.03	4.4136	3.2299
#1	.00005	.00069	-.00054	.02667	.18563	.00276	1.4626	.08626
#2	.00029	-.00017	.00042	.03166	.01192	-.00066	1.4861	.08273
#3	-.00048	-.00010	-.00042	.02972	.08953	-.00352	1.3663	.08818
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00023	3.8750	.00064	.00454	.00264	-.00168	.00429	.37174
Stddev	.00028	.0249	.00026	.00324	.00293	.00204	.00272	.00091
%RSD	124.52	.64140	39.830	71.430	111.04	121.37	63.452	.24386
#1	-.00008	3.8959	.00063	.00081	.00037	-.00400	.00736	.37278
#2	.00028	3.8475	.00040	.00665	.00594	-.00087	.00338	.37129
#3	.00048	3.8816	.00091	.00616	.00160	-.00017	.00214	.37115
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115501 Acquired: 6/25/2016 17:06:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00004**	**.01613**	**.00436**	**.00049**	**-.00004**	**.00087**	**.00142**
Stddev	.00092	.00051	.00120	.00199	.00066	.00012	.00191
%RSD	2230.2	3.1791	27.456	410.68	1766.6	13.339	134.89
#1	.00087	.01607	.00571	.00178	-.00027	.00083	.00247
#2	-.00002	.01565	.00341	-.00181	.00071	.00100	-.00079
#3	-.00097	.01667	.00397	.00149	-.00055	.00078	.00257
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9938.9**	**60110.**	**3041.6**
Stddev	14.3	99.	8.4
%RSD	.14384	.16440	.27454
#1	9922.4	60057.	3050.1
#2	9946.9	60224.	3033.5
#3	9947.4	60049.	3041.2

Approved: June 27, 2016

Sample Name: L1606115502 Acquired: 6/25/2016 17:10:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00066	-.00453	-.00033	.00989	.01794	.00013	2.4889	.00011
Stddev	.00144	.01600	.00099	.00264	.00078	.00009	.0280	.00014
%RSD	216.56	353.18	295.78	26.673	4.3403	71.235	1.1261	129.84
#1	.00231	.00655	-.00001	.00702	.01871	.00024	2.4937	.00027
#2	.00001	-.02287	-.00144	.01220	.01715	.00010	2.5143	-.00000
#3	-.00033	.00273	.00045	.01047	.01798	.00006	2.4588	.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00008	-.00044	.00104	.24605	-.07520	-.00217	1.6643	.02363
Stddev	.00022	.00012	.00019	.01179	.12175	.00286	.0624	.00224
%RSD	272.66	27.286	18.718	4.7938	161.89	131.42	3.7522	9.4751
#1	.00027	-.00041	.00090	.23638	.00559	-.00035	1.5996	.02553
#2	.00015	-.00033	.00126	.24256	-.01597	-.00070	1.6689	.02116
#3	-.00017	-.00057	.00095	.25919	-.21524	-.00546	1.7243	.02420
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00021	2.8989	.00016	.00365	.00260	-.00217	.00870	.42450
Stddev	.00065	.0101	.00120	.00396	.00147	.00150	.00507	.00601
%RSD	314.77	.35003	759.93	108.45	56.477	68.876	58.304	1.4160
#1	-.00082	2.8946	-.00099	.00772	.00422	-.00382	.01102	.41775
#2	-.00027	2.9105	.00140	-.00019	.00221	-.00091	.01219	.42650
#3	.00047	2.8917	.00006	.00342	.00137	-.00178	.00288	.42926
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115502 Acquired: 6/25/2016 17:10:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00038**	**.02139**	**.00454**	**-.00050**	**.00021**	**.00050**	**.00118**
Stddev	.00016	.00022	.00843	.00114	.00027	.00017	.00137
%RSD	41.986	1.0329	185.67	229.51	129.67	34.277	116.06
#1	.00030	.02116	.00016	.00076	.00041	.00055	-.00040
#2	.00027	.02139	.01426	-.00146	.00031	.00031	.00208
#3	.00056	.02161	-.00080	-.00078	-.00010	.00064	.00187
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9945.8**	**60248.**	**3055.3**
Stddev	9.2	94.	11.4
%RSD	.09217	.15541	.37444
#1	9945.9	60314.	3061.2
#2	9954.9	60289.	3042.1
#3	9936.5	60141.	3062.6

Approved: June 27, 2016

Sample Name: L1606115503 Acquired: 6/25/2016 17:14:07 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00200	-.00135	-.00104	.00757	.00737	.00014	2.3997	-.00010
Stddev	.00192	.00661	.00046	.00116	.00061	.00003	.0309	.00022
%RSD	96.141	489.89	44.488	15.290	8.2167	21.217	1.2897	218.06
#1	.00406	-.00575	-.00051	.00624	.00716	.00012	2.3659	-.00020
#2	.00025	-.00456	-.00131	.00835	.00689	.00017	2.4268	-.00025
#3	.00170	.00625	-.00131	.00813	.00805	.00012	2.4063	.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00001	-.00044	-.00019	.10432	-.04481	.00004	.91519	.02191
Stddev	.00035	.00082	.00101	.01415	.09851	.00279	.09371	.00121
%RSD	3296.1	188.98	541.26	13.565	219.83	6329.0	10.240	5.5295
#1	-.00033	.00027	.00027	.10406	-.13456	.00163	.99135	.02326
#2	-.00007	-.00023	-.00134	.09030	-.06048	.00168	.94369	.02093
#3	.00037	-.00134	.00052	.11860	.06059	-.00317	.81054	.02153
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00009	2.4900	.00017	.00956	.00115	-.00111	.00305	.32330
Stddev	.00005	.0408	.00057	.00296	.00162	.00041	.00349	.00479
%RSD	56.574	1.6398	345.24	30.912	140.85	37.279	114.64	1.4819
#1	-.00011	2.4429	.00078	.00752	.00187	-.00066	-.00055	.31801
#2	-.00013	2.5161	.00008	.00822	-.00070	-.00119	.00642	.32454
#3	-.00003	2.5109	-.00036	.01295	.00229	-.00148	.00327	.32735
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115503 Acquired: 6/25/2016 17:14:07 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00009**	**.01387**	**.00305**	**-.00065**	**.00052**	**.00084**	**.00071**
Stddev	.00081	.00016	.00789	.00174	.00073	.00004	.00140
%RSD	953.07	1.1448	258.94	265.86	139.13	5.0666	198.28
#1	-.00013	.01403	.00198	-.00023	.00032	.00082	.00133
#2	-.00088	.01386	-.00426	.00083	-.00008	.00082	.00169
#3	.00075	.01371	.01142	-.00257	.00133	.00089	-.00090
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9910.7**	**60184.**	**3019.4**
Stddev	10.3	175.	9.0
%RSD	.10381	.29035	.29953
#1	9910.7	60334.	3026.1
#2	9921.0	60225.	3023.1
#3	9900.4	59992.	3009.1

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 17:18:04 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41531**	**10.447**	**.40916**	**.50533**	**1.0238**	**.05217**	**10.225**
Stddev	.00091	.032	.00132	.00054	.0034	.00011	.030
%RSD	.21998	.31066	.32348	.10618	.32870	.20293	.29201
#1	.41613	10.474	.41061	.50566	1.0229	.05229	10.191
#2	.41432	10.411	.40886	.50561	1.0276	.05213	10.247
#3	.41546	10.455	.40801	.50471	1.0211	.05209	10.235
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05225**	**.20370**	**.51712**	**.52549**	**4.0444**	**49.392**	**1.0301**
Stddev	.00031	.00034	.00175	.00197	.0184	.117	.0048
%RSD	.59442	.16603	.33927	.37570	.45604	.23752	.46659
#1	.05229	.20396	.51906	.52418	4.0394	49.465	1.0330
#2	.05254	.20382	.51666	.52776	4.0648	49.455	1.0327
#3	.05192	.20332	.51564	.52452	4.0290	49.257	1.0245
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.7951**	**.51327**	**1.0423**	**50.236**	**.51912**	**10.208**	**.51997**
Stddev	.0472	.00504	.0084	.169	.00170	.017	.00372
%RSD	.48165	.98200	.80990	.33677	.32826	.16636	.71476
#1	9.7523	.51223	1.0511	50.262	.52046	10.219	.52035
#2	9.7875	.51875	1.0343	50.391	.51968	10.189	.51607
#3	9.8457	.50883	1.0415	50.056	.51720	10.217	.52348
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 17:18:04 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2238**	**.40709**	**5.2468**	**1.0260**	**1.0230**	**1.0304**	**.52875**
Stddev	.0084	.00511	.0098	.0012	.0023	.0127	.00327
%RSD	.68659	1.2545	.18586	.11540	.22459	1.2311	.61859
#1	1.2260	.40344	5.2580	1.0271	1.0231	1.0165	.53252
#2	1.2308	.40491	5.2428	1.0262	1.0252	1.0415	.52673
#3	1.2145	.41292	5.2398	1.0248	1.0206	1.0331	.52699
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0570**	**1.0310**	F **3.0075**
Stddev	.0056	.0001	.0207
%RSD	.53103	.00939	.68856
#1	1.0633	1.0311	3.0004
#2	1.0553	1.0309	3.0308
#3	1.0525	1.0310	2.9913
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9557.2**	**57623.**	**2979.5**
Stddev	9.7	193.	16.3
%RSD	.10109	.33465	.54600
#1	9546.7	57405.	2979.8
#2	9565.7	57771.	2963.0
#3	9559.4	57693.	2995.6

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 17:21:42 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**-.00563**	**.00200**	**-.00100**	**.00189**	**.00009**	**-.01127**	**-.00004**
Stddev	.00104	.00334	.00233	.00184	.00120	.00004	.03123	.00010
%RSD	1246.5	59.297	116.54	184.11	63.245	40.940	277.20	226.69
#1	-.00074	-.00335	.00466	-.00138	.00302	.00011	-.03377	.00004
#2	-.00027	-.00407	.00108	-.00263	.00202	.00005	.02439	-.00015
#3	.00126	-.00946	.00027	.00100	.00064	.00010	-.02442	-.00003
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00016**	**.00048**	**.00089**	**.00955**	**-.03243**	**.00071**	**-.01124**	**-.00013**
Stddev	.00017	.00075	.00043	.01226	.03763	.00268	.07316	.00129
%RSD	103.61	157.05	48.720	128.38	116.03	378.57	650.98	975.30
#1	.00002	.00095	.00059	-.00410	-.06003	-.00175	-.02665	-.00039
#2	.00035	.00087	.00069	.01962	.01043	.00031	-.07547	-.00128
#3	.00012	-.00039	.00138	.01314	-.04771	.00356	.06840	.00127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00323**	**.02898**	**-.00046**	**.00432**	**.00328**	**.00454**	**.00410**	**-.00425**
Stddev	.00029	.02008	.00095	.00359	.00221	.00123	.00133	.00064
%RSD	9.1168	69.292	208.78	82.910	67.490	26.995	32.528	15.135
#1	.00298	.02490	-.00038	.00845	.00352	.00323	.00275	-.00470
#2	.00356	.01126	-.00144	.00261	.00095	.00566	.00541	-.00351
#3	.00316	.05079	.00046	.00192	.00536	.00474	.00413	-.00454
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 17:21:42 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00068	.00021	.00375	-.00110	.00108	-.00009	.00393
Stddev	.00059	.00011	.00446	.00333	.00027	.00003	.00087
%RSD	86.694	51.296	119.05	302.86	25.045	36.058	22.188
#1	.00135	.00012	.00567	.00135	.00136	-.00006	.00491
#2	.00024	.00018	-.00135	.00024	.00107	-.00012	.00363
#3	.00044	.00033	.00692	-.00489	.00082	-.00008	.00324
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9706.7	59138.	3007.6
Stddev	9.6	122.	14.5
%RSD	.09901	.20648	.48341
#1	9697.1	58997.	2992.2
#2	9716.3	59217.	3009.4
#3	9706.6	59200.	3021.1

Approved: June 27, 2016

Sample Name: LLCCV Acquired: 6/25/2016 17:25:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00682	.16326	.00787	.08110	.01015	.00172	.37504	.00080
Stddev	.00230	.01176	.00393	.00058	.00060	.00002	.03274	.00014
%RSD	33.789	7.2031	49.909	.72089	5.8834	1.2548	8.7301	17.457
#1	.00456	.16635	.01239	.08173	.01084	.00173	.33805	.00070
#2	.00916	.17317	.00587	.08058	.00985	.00173	.40031	.00075
#3	.00674	.15026	.00534	.08099	.00976	.00169	.38675	.00096
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00446	.00390	.00484	.09497	.71355	.08241	.33888	.00741
Stddev	.00036	.00087	.00066	.01610	.00812	.00379	.00870	.00176
%RSD	8.1660	22.291	13.622	16.951	1.1379	4.5949	2.5673	23.707
#1	.00488	.00490	.00494	.08463	.70937	.08497	.33660	.00920
#2	.00426	.00332	.00413	.08675	.70838	.08419	.34850	.00734
#3	.00424	.00348	.00544	.11352	.72291	.07806	.33155	.00569
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00855	.40833	.01678	.78916	.01199	.08269	.02089	.77549
Stddev	.00007	.02661	.00076	.00972	.00016	.00111	.00270	.00482
%RSD	.82014	6.5175	4.5074	1.2317	1.3733	1.3390	12.938	.62171
#1	.00856	.42291	.01686	.80033	.01215	.08397	.01985	.78057
#2	.00847	.42447	.01749	.78261	.01182	.08216	.01886	.77494
#3	.00861	.37762	.01598	.78454	.01200	.08195	.02396	.77097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LLCCV Acquired: 6/25/2016 17:25:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.40481	.04214	.02257	.15413	.00994	.01699	.13476
Stddev	.00262	.00003	.00404	.00144	.00108	.00007	.00396
%RSD	.64734	.07891	17.895	.93373	10.832	.40613	2.9378
#1	.40763	.04217	.02259	.15502	.01117	.01705	.13268
#2	.40435	.04210	.02661	.15491	.00942	.01692	.13226
#3	.40245	.04214	.01853	.15247	.00922	.01700	.13932
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9757.6	59331.	3007.6
Stddev	3.7	156.	34.2
%RSD	.03744	.26375	1.1357
#1	9754.3	59400.	3008.1
#2	9757.0	59152.	3041.5
#3	9761.5	59441.	2973.2

Approved: June 27, 2016

Sample Name: PBW Y5 Acquired: 6/25/2016 17:29:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00185	.00116	-.00112	-.00084	.00120	.00011	-.01961	-.00006
Stddev	.00045	.01041	.00198	.00240	.00080	.00010	.05200	.00016
%RSD	24.231	897.41	176.31	284.10	67.234	89.198	265.17	288.76
#1	.00153	-.00604	-.00020	.00070	.00046	.00020	.01383	-.00024
#2	.00236	-.00358	.00023	-.00360	.00107	.00014	-.07952	.00007
#3	.00166	.01310	-.00339	.00037	.00206	.00000	.00686	.00000
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00010	-.00018	.00105	.03412	-.07407	.00236	.02694	.00162
Stddev	.00024	.00064	.00070	.02542	.08098	.00211	.02742	.00224
%RSD	248.33	362.92	67.168	74.514	109.32	89.477	101.81	137.94
#1	.00030	-.00083	.00024	.00498	-.16192	.00417	.01237	.00092
#2	.00016	-.00014	.00146	.05181	-.05788	.00285	.00987	.00413
#3	-.00017	.00044	.00145	.04557	-.00242	.00004	.05857	-.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00014	.01025	-.00077	.00507	.00386	-.00192	-.00109	.00437
Stddev	.00004	.01479	.00185	.00580	.00053	.00084	.00245	.00068
%RSD	28.058	144.27	239.74	114.27	13.817	43.798	224.68	15.497
#1	-.00010	.02026	-.00256	.00797	.00348	-.00273	.00112	.00408
#2	-.00015	-.00674	.00113	.00885	.00447	-.00198	-.00067	.00389
#3	-.00017	.01724	-.00088	-.00160	.00363	-.00105	-.00372	.00515
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: PBW Y5 Acquired: 6/25/2016 17:29:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00077	.00044	.00538	-.00184	.00071	.00062	.00198
Stddev	.00057	.00040	.00350	.00250	.00039	.00016	.00115
%RSD	73.146	89.773	65.128	135.94	54.397	26.345	58.223
#1	.00143	.00035	.00942	-.00427	.00062	.00079	.00279
#2	.00040	.00088	.00331	.00073	.00114	.00048	.00066
#3	.00049	.00010	.00340	-.00198	.00038	.00058	.00249
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9880.3	60109.	3027.1
Stddev	31.5	73.	6.6
%RSD	.31885	.12179	.21761
#1	9847.3	60149.	3030.6
#2	9910.1	60154.	3019.5
#3	9883.5	60025.	3031.2

Approved: June 27, 2016

Sample Name: LCSW Y5 Acquired: 6/25/2016 17:33:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.18967**	**4.7620**	**.18828**	**.90013**	**.47555**	**.02395**	**4.7950**	**.02373**
Stddev	.00250	.0081	.00104	.00471	.00273	.00006	.0213	.00018
%RSD	1.3192	.17078	.55148	.52284	.57347	.24848	.44480	.74408
#1	.18907	4.7578	.18718	.89893	.47827	.02388	4.7819	.02392
#2	.19242	4.7714	.18841	.90532	.47282	.02399	4.8196	.02370
#3	.18753	4.7569	.18924	.89614	.47557	.02397	4.7835	.02357
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09489**	**.23803**	**.24292**	**1.9123**	**22.817**	**.48875**	**4.5669**	**.24181**
Stddev	.00039	.00120	.00156	.0298	.036	.00284	.0465	.00007
%RSD	.41346	.50457	.64104	1.5597	.15629	.58110	1.0173	.02751
#1	.09520	.23794	.24291	1.9317	22.780	.48603	4.5750	.24188
#2	.09501	.23927	.24447	1.8779	22.852	.49170	4.5170	.24174
#3	.09445	.23687	.24136	1.9272	22.818	.48851	4.6088	.24180
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49027**	**23.490**	**.24091**	**4.5553**	**.24063**	**.56646**	**.18382**	**2.3525**
Stddev	.00145	.113	.00040	.0084	.00352	.00190	.00239	.0014
%RSD	.29550	.48187	.16805	.18553	1.4632	.33509	1.2995	.06062
#1	.48963	23.608	.24065	4.5528	.24014	.56531	.18109	2.3541
#2	.49193	23.382	.24137	4.5647	.24437	.56542	.18555	2.3519
#3	.48925	23.480	.24069	4.5484	.23738	.56865	.18482	2.3515
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LCSW Y5 Acquired: 6/25/2016 17:33:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47085**	**.48454**	**.47923**	**.24546**	**.48968**	**.47202**	**.01198**
Stddev	.00082	.00070	.00116	.00117	.00078	.00042	.00247
%RSD	.17390	.14504	.24256	.47677	.15993	.08855	20.639
#1	.47170	.48509	.47789	.24652	.48877	.47178	.00945
#2	.47079	.48375	.47984	.24421	.49011	.47251	.01208
#3	.47006	.48479	.47995	.24566	.49015	.47178	.01439
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9888.6**	**60001.**	**3090.1**
Stddev	19.1	149.	26.5
%RSD	.19299	.24892	.85667
#1	9894.9	60052.	3063.1
#2	9867.1	59834.	3091.3
#3	9903.7	60119.	3116.0

Approved: June 27, 2016

Sample Name: F BLANK Acquired: 6/25/2016 17:37:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00166	-.00148	.00090	.00227	.00163	.00011	-.00468	.00008
Stddev	.00130	.00490	.00188	.00068	.00037	.00003	.02546	.00022
%RSD	78.305	330.55	208.29	29.974	22.968	25.204	544.18	269.24
#1	.00033	-.00239	.00275	.00286	.00197	.00009	-.03398	.00015
#2	.00173	.00381	-.00100	.00243	.00123	.00014	.01206	.00026
#3	.00293	-.00586	.00096	.00153	.00167	.00010	.00788	-.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00010	.00058	.00129	.04476	.01060	-.00161	-.05710	.00063
Stddev	.00029	.00122	.00029	.02153	.02075	.00622	.06303	.00318
%RSD	276.09	211.55	22.793	48.099	195.81	387.03	110.39	502.59
#1	-.00039	.00184	.00121	.03412	-.00619	.00062	-.09553	.00029
#2	.00018	-.00059	.00161	.06953	.03381	.00319	-.09140	.00397
#3	-.00010	.00048	.00104	.03061	.00418	-.00863	.01564	-.00236
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00056	.05677	-.00011	.00316	.00576	-.00381	.00560	-.00010
Stddev	.00040	.03222	.00050	.00162	.00079	.00329	.00231	.00141
%RSD	72.047	56.761	453.37	51.247	13.705	86.399	41.256	1388.8
#1	.00085	.07680	-.00069	.00292	.00611	-.00013	.00434	-.00092
#2	.00073	.01960	.00014	.00168	.00631	-.00483	.00420	-.00091
#3	.00010	.07391	.00021	.00489	.00485	-.00647	.00827	.00153
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: F BLANK Acquired: 6/25/2016 17:37:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00007	.00021	-.00396	-.00081	.00112	.00164	.00166
Stddev	.00084	.00011	.00377	.00288	.00053	.00013	.00011
%RSD	1212.1	52.839	95.183	357.63	47.087	8.1028	6.9152
#1	-.00082	.00025	-.00813	-.00413	.00144	.00151	.00159
#2	.00085	.00029	-.00300	.00065	.00141	.00165	.00179
#3	.00018	.00008	-.00077	.00106	.00051	.00177	.00159
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10015.	61314.	3090.8
Stddev	16.	177.	23.0
%RSD	.15522	.28891	.74279
#1	10012.	61143.	3064.6
#2	10032.	61303.	3107.1
#3	10001.	61497.	3100.8

Approved: June 27, 2016

Sample Name: L1606098602 Acquired: 6/25/2016 17:41:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00088	-.00042	-.00040	.09099	.02315	.00010	31.081	-.00006
Stddev	.00141	.00402	.00168	.00026	.00060	.00002	.292	.00002
%RSD	160.55	951.46	416.45	.28845	2.5949	20.817	.93935	37.551
#1	.00197	.00339	-.00222	.09093	.02306	.00008	31.173	-.00006
#2	-.00072	-.00463	.00109	.09127	.02259	.00012	30.754	-.00004
#3	.00139	-.00003	-.00008	.09076	.02379	.00010	31.315	-.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00024	.00059	.00094	.04555	.98402	.01199	5.7014	.03480
Stddev	.00048	.00016	.00041	.01573	.08850	.00357	.0718	.00212
%RSD	197.55	26.299	43.232	34.544	8.9935	29.756	1.2595	6.0810
#1	.00072	.00042	.00054	.02747	.88485	.01459	5.6202	.03718
#2	-.00024	.00064	.00135	.05303	1.0122	.01345	5.7566	.03314
#3	.00026	.00072	.00092	.05614	1.0550	.00792	5.7274	.03406
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00142	119.95	.00080	.00488	.00205	-.00278	-.00166	3.2347
Stddev	.00007	.78	.00132	.00386	.00179	.00227	.00673	.0109
%RSD	4.5853	.65032	165.26	79.044	87.096	81.844	406.14	.33815
#1	.00137	120.20	.00201	.00062	.00137	-.00037	-.00009	3.2461
#2	.00149	119.07	.00100	.00813	.00409	-.00489	.00415	3.2336
#3	.00140	120.57	-.00061	.00590	.00071	-.00307	-.00903	3.2243
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098602 Acquired: 6/25/2016 17:41:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00169	.54147	-.00161	-.00445	.00062	.00187	.00166
Stddev	.00060	.00457	.00418	.00188	.00027	.00011	.00091
%RSD	35.360	.84355	259.35	42.200	43.191	5.8148	54.635
#1	.00106	.54395	-.00040	-.00269	.00035	.00198	.00205
#2	.00225	.53620	.00182	-.00642	.00063	.00186	.00062
#3	.00175	.54425	-.00626	-.00422	.00089	.00176	.00230
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9793.0	59457.	3063.6
Stddev	12.2	310.	25.9
%RSD	.12488	.52191	.84399
#1	9783.4	59142.	3057.4
#2	9788.8	59467.	3092.0
#3	9806.7	59762.	3041.4

Approved: June 27, 2016

Sample Name: L1606098604 Acquired: 6/25/2016 17:45:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00188**	**.00192**	**.00240**	**.01933**	**.01606**	**.00017**	**258.27**
Stddev	.00168	.00663	.00025	.00142	.00064	.00002	.74
%RSD	89.608	344.23	10.527	7.3208	4.0034	12.550	.28551
#1	.00265	-.00052	.00267	.01828	.01569	.00019	258.08
#2	.00303	-.00313	.00217	.01877	.01568	.00016	257.65
#3	-.00005	.00943	.00237	.02094	.01680	.00015	259.09
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00045**	**.00038**	**.00085**	**.01097**	**.02808**	**1.7243**	**.01739**
Stddev	.00023	.00042	.00027	.00107	.01426	.0758	.00077
%RSD	50.915	110.30	31.401	9.7605	50.795	4.3949	4.4280
#1	.00057	.00083	.00065	.01199	.02723	1.7078	.01656
#2	.00019	.00033	.00115	.00986	.01426	1.6581	.01808
#3	.00061	-.00001	.00075	.01107	.04275	1.8069	.01753
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**67.436**	**1.3917**	**.00099**	**14.380**	**.00387**	**.00116**	**.00177**
Stddev	.211	.0043	.00004	.031	.00051	.00485	.00143
%RSD	.31235	.30797	4.3492	.21337	13.178	419.65	80.619
#1	67.195	1.3899	.00099	14.411	.00414	.00253	.00287
#2	67.533	1.3886	.00103	14.350	.00328	-.00423	.00229
#3	67.581	1.3966	.00095	14.377	.00419	.00517	.00016
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: L1606098604 Acquired: 6/25/2016 17:45:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00130**	**.00146**	**3.7576**	**.00044**	**1.2471**	F **-.03982**	**-.00203**
Stddev	.00495	.00434	.0110	.00062	.0034	.00406	.00082
%RSD	381.66	297.39	.29276	142.20	.26857	10.186	40.503
#1	-.00325	.00313	3.7635	.00113	1.2483	-.04298	-.00122
#2	.00433	.00472	3.7644	.00025	1.2433	-.03524	-.00200
#3	-.00498	-.00347	3.7449	-.00007	1.2496	-.04122	-.00286
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00050**	**.03051**	**.00151**
Stddev	.00029	.00016	.00074
%RSD	57.554	.53953	48.929
#1	-.00084	.03044	.00164
#2	-.00033	.03070	.00216
#3	-.00034	.03040	.00071
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9610.5**	**58897.**	**3044.6**
Stddev	3.8	97.	5.6
%RSD	.03917	.16551	.18408
#1	9614.6	58944.	3043.4
#2	9607.2	58961.	3050.7
#3	9609.7	58785.	3039.7

Approved: June 27, 2016

Sample Name: L1606098606 Acquired: 6/25/2016 17:48:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00223	.00972	-.00038	.01373	.06613	.00026	77.440	.00018
Stddev	.00105	.00309	.00132	.00205	.00021	.00009	.075	.00021
%RSD	47.129	31.833	346.72	14.955	.32162	33.087	.09643	116.74
#1	.00235	.00691	-.00086	.01602	.06623	.00027	77.417	.00042
#2	.00112	.00920	-.00139	.01205	.06589	.00034	77.379	.00001
#3	.00321	.01303	.00111	.01311	.06628	.00017	77.523	.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00039	.00108	.01311	.03469	1.0666	.00528	22.814	.00226
Stddev	.00015	.00032	.00053	.01962	.0266	.00092	.288	.00095
%RSD	37.987	29.419	4.0055	56.546	2.4929	17.478	1.2638	41.753
#1	.00056	.00082	.01330	.03776	1.0701	.00541	22.650	.00309
#2	.00027	.00143	.01252	.05259	1.0384	.00430	22.645	.00123
#3	.00036	.00097	.01352	.01372	1.0912	.00613	23.147	.00247
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00038	16.230	.00086	.01073	.00099	-.00405	.00160	2.3863
Stddev	.00023	.056	.00046	.00069	.00181	.00317	.00143	.0169
%RSD	59.834	.34372	53.276	6.4188	182.38	78.406	89.367	.70593
#1	.00032	16.197	.00073	.01029	.00246	-.00129	-.00004	2.4028
#2	.00063	16.198	.00137	.01153	-.00103	-.00752	.00258	2.3871
#3	.00019	16.294	.00048	.01038	.00155	-.00333	.00227	2.3691
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098606 Acquired: 6/25/2016 17:48:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00004**	**.39726**	**-.00928**	**-.00135**	**-.00044**	**.03812**	**.00202**
Stddev	.00081	.00077	.00327	.00251	.00044	.00033	.00118
%RSD	1967.6	.19375	35.277	185.96	101.24	.85896	58.324
#1	.00088	.39645	-.00551	-.00111	-.00048	.03844	.00108
#2	-.00002	.39799	-.01103	.00103	.00002	.03814	.00334
#3	-.00074	.39735	-.01131	-.00397	-.00086	.03779	.00164
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9875.4**	**60112.**	**3085.9**
Stddev	17.7	60.	4.4
%RSD	.17969	.09972	.14239
#1	9863.6	60084.	3084.9
#2	9866.7	60181.	3090.7
#3	9895.8	60072.	3082.1

Approved: June 27, 2016

Sample Name: L1606098608 Acquired: 6/25/2016 17:52:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00010	.00115	.00027	.01585	.03262	.00011	84.217	.00013
Stddev	.00235	.00513	.00111	.00136	.00141	.00004	.314	.00017
%RSD	2475.5	446.86	409.02	8.5842	4.3250	37.715	.37277	125.68
#1	-.00262	.00404	-.00099	.01740	.03376	.00010	84.484	.00019
#2	.00142	-.00477	.00110	.01531	.03104	.00016	84.297	.00026
#3	.00149	.00417	.00071	.01485	.03305	.00008	83.871	-.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00015	.00177	.00567	.03434	1.2401	.01415	23.228	.02727
Stddev	.00046	.00068	.00058	.01787	.0425	.00095	.227	.00106
%RSD	299.90	38.444	10.195	52.028	3.4256	6.6956	.97712	3.8827
#1	-.00049	.00241	.00506	.05413	1.2665	.01352	23.122	.02850
#2	.00037	.00184	.00573	.01939	1.1911	.01369	23.489	.02663
#3	-.00034	.00105	.00621	.02950	1.2628	.01524	23.073	.02669
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00074	12.660	.00033	.00413	.00179	-.00334	-.00002	3.4869
Stddev	.00017	.055	.00044	.00392	.00093	.00378	.00639	.0225
%RSD	23.632	.43231	134.67	94.801	51.837	113.14	36088.	.64396
#1	.00093	12.716	.00084	.00852	.00134	-.00400	-.00400	3.4968
#2	.00069	12.657	.00002	.00099	.00117	-.00675	-.00341	3.5027
#3	.00059	12.607	.00013	.00288	.00285	.00073	.00736	3.4612
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098608 Acquired: 6/25/2016 17:52:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00019**	**.83876**	**-.00998**	**-.00331**	**.00001**	**.01681**	**.00058**
Stddev	.00022	.00442	.00442	.00242	.00077	.00018	.00259
%RSD	112.09	.52651	44.282	73.168	6824.2	1.0650	446.63
#1	-.00039	.84240	-.01091	-.00183	-.00011	.01673	-.00103
#2	-.00022	.84004	-.00517	-.00610	-.00069	.01701	.00357
#3	.00004	.83385	-.01387	-.00199	.00083	.01668	-.00080
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9863.7**	**59966.**	**3069.4**
Stddev	33.8	107.	16.4
%RSD	.34299	.17793	.53396
#1	9879.3	59844.	3052.8
#2	9824.9	60042.	3069.7
#3	9886.9	60012.	3085.6

Approved: June 27, 2016

Sample Name: L1606098608PS Acquired: 6/25/2016 17:56:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG574216-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.19012	4.7962	.18985	.92761	.50579	.02416	80.589	.02376
Stddev	.00115	.0181	.00324	.00188	.00153	.00010	.509	.00031
%RSD	.60416	.37710	1.7085	.20237	.30247	.42369	.63122	1.3080
#1	.18881	4.7997	.18690	.92863	.50534	.02415	80.494	.02389
#2	.19096	4.8123	.19332	.92876	.50750	.02427	81.138	.02341
#3	.19058	4.7766	.18933	.92544	.50454	.02406	80.134	.02399
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09205	.23624	.24131	1.9065	24.291	.50517	25.434	.26103
Stddev	.00018	.00146	.00130	.0143	.039	.00496	.150	.00626
%RSD	.19260	.61641	.53797	.74911	.16017	.98165	.59175	2.3977
#1	.09210	.23659	.24071	1.9177	24.287	.49968	25.563	.26590
#2	.09186	.23464	.24280	1.9113	24.254	.50647	25.268	.26320
#3	.09221	.23749	.24041	1.8904	24.332	.50934	25.470	.25397
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.49345	34.822	.23345	4.6884	.23609	.56850	.18268	5.5114
Stddev	.00019	.169	.00142	.0142	.00258	.00539	.00604	.0369
%RSD	.03860	.48533	.60979	.30359	1.0940	.94754	3.3047	.66920
#1	.49323	34.790	.23475	4.6782	.23789	.57472	.18356	5.4689
#2	.49357	35.005	.23367	4.6823	.23725	.56528	.18824	5.5341
#3	.49354	34.671	.23193	4.7047	.23313	.56551	.17626	5.5313
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098608PS Acquired: 6/25/2016 17:56:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG574216-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47094**	**1.2232**	**.47334**	**.23988**	**.49331**	**.47631**	**.00522**
Stddev	.00189	.0075	.00962	.00117	.00125	.00347	.00156
%RSD	.40190	.61543	2.0323	.48785	.25347	.72879	29.825
#1	.47068	1.2227	.46319	.24082	.49476	.47230	.00351
#2	.47295	1.2310	.48233	.23857	.49254	.47818	.00558
#3	.46920	1.2160	.47451	.24025	.49265	.47845	.00657
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9831.9**	**59526.**	**3075.8**
Stddev	15.2	194.	19.0
%RSD	.15421	.32611	.61679
#1	9829.8	59604.	3066.0
#2	9817.9	59305.	3063.8
#3	9848.0	59669.	3097.7

Approved: June 27, 2016

Sample Name: L1606098608SDL Acquired: 6/25/2016 18:00:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG574216-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00260	-.00514	.00187	.00508	.00868	.00012	17.605	-.00015
Stddev	.00156	.01146	.00424	.00200	.00063	.00008	.142	.00010
%RSD	60.035	222.87	226.36	39.430	7.2120	66.284	.80739	66.159
#1	.00437	.00780	.00118	.00739	.00923	.00013	17.451	-.00025
#2	.00143	-.00924	.00642	.00396	.00881	.00020	17.731	-.00007
#3	.00199	-.01399	-.00198	.00389	.00800	.00004	17.633	-.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00018	.00006	.00193	.02399	.21059	-.00038	4.9921	.00509
Stddev	.00017	.00044	.00078	.00583	.02318	.00216	.0797	.00111
%RSD	97.856	772.57	40.443	24.318	11.008	572.98	1.5970	21.792
#1	-.00004	-.00041	.00282	.01859	.23319	.00193	5.0036	.00623
#2	-.00037	.00046	.00138	.03017	.18686	-.00072	4.9072	.00402
#3	-.00011	.00012	.00158	.02319	.21171	-.00234	5.0654	.00500
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00034	2.6677	-.00071	.00513	.00009	-.00286	-.00058	.72991
Stddev	.00017	.0114	.00035	.00552	.00151	.00279	.00302	.00273
%RSD	51.097	.42653	49.400	107.62	1685.9	97.489	522.92	.37384
#1	.00017	2.6710	-.00044	.00122	.00180	.00001	.00149	.72939
#2	.00033	2.6550	-.00058	.01144	-.00102	-.00304	-.00405	.73285
#3	.00051	2.6770	-.00110	.00273	-.00051	-.00555	.00083	.72747
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098608SDL Acquired: 6/25/2016 18:00:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG574216-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00006**	**.17684**	**.00226**	**.00019**	**.00052**	**.00413**	**.00011**
Stddev	.00027	.00102	.00232	.00368	.00057	.00014	.00068
%RSD	464.39	.57572	102.44	1951.0	110.67	3.3962	623.50
#1	-.00013	.17748	.00403	.00128	.00008	.00397	.00059
#2	.00024	.17567	-.00036	.00320	.00031	.00420	.00041
#3	-.00028	.17737	.00312	-.00392	.00117	.00423	-.00067
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9896.0**	**60261.**	**3042.3**
Stddev	25.0	81.	28.3
%RSD	.25242	.13391	.92921
#1	9907.7	60289.	3040.8
#2	9913.0	60324.	3071.3
#3	9867.3	60170.	3014.9

Approved: June 27, 2016

Sample Name: L1606098608SDL Acquired: 6/25/2016 18:04:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG574216-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00175	-.00807	-.00063	-.00078	.00182	.00014	3.3399	-.00019
Stddev	.00147	.00382	.00127	.00273	.00016	.00003	.0946	.00023
%RSD	83.969	47.279	200.74	348.35	8.5330	23.588	2.8330	121.57
#1	.00097	-.00874	.00010	.00038	.00196	.00018	3.2721	-.00045
#2	.00084	-.00396	.00010	.00117	.00166	.00011	3.4480	-.00012
#3	.00344	-.01151	-.00210	-.00390	.00185	.00014	3.2995	-.00000
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00003	-.00037	.00107	.01769	-.08284	.00119	.94569	.00052
Stddev	.00047	.00133	.00130	.01702	.09737	.00358	.05932	.00055
%RSD	1445.4	360.33	120.72	96.212	117.53	302.26	6.2728	105.20
#1	.00035	-.00086	.00167	-.00174	-.18470	.00532	.88043	.00057
#2	-.00050	-.00138	-.00041	.02995	-.07314	-.00085	.96028	.00105
#3	.00025	.00113	.00196	.02486	.00931	-.00092	.99635	-.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00012	.49987	-.00021	.00755	.00303	-.00541	.00850	.13812
Stddev	.00032	.04592	.00020	.00386	.00188	.00314	.00116	.00966
%RSD	265.16	9.1856	96.762	51.054	62.049	58.070	13.660	6.9939
#1	.00001	.44687	-.00020	.01173	.00086	-.00531	.00838	.12921
#2	-.00049	.52761	-.00041	.00680	.00410	-.00232	.00740	.13676
#3	.00011	.52513	-.00001	.00413	.00413	-.00859	.00972	.14839
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098608SDL Acquired: 6/25/2016 18:04:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG574216-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00039**	**.03350**	**.00349**	**.00073**	**.00046**	**.00106**	**.00161**
Stddev	.00032	.00105	.00672	.00069	.00093	.00001	.00112
%RSD	80.289	3.1425	192.33	93.894	202.19	.93440	69.701
#1	.00063	.03282	.00553	-.00006	-.00031	.00105	.00058
#2	.00051	.03471	.00896	.00114	.00020	.00107	.00146
#3	.00004	.03296	-.00401	.00113	.00149	.00107	.00281
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10400.**	**61068.**	**3085.8**
Stddev	302.	416.	38.3
%RSD	2.9005	.68044	1.2414
#1	10744.	61402.	3085.7
#2	10277.	61200.	3047.5
#3	10180.	60603.	3124.1

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 18:08:15 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.40600	10.196	.39897	.49353	1.0097	.05111	10.072
Stddev	.00134	.005	.00384	.00145	.0055	.00016	.108
%RSD	.32964	.04609	.96241	.29468	.54041	.30777	1.0683
#1	.40454	10.197	.40224	.49289	1.0058	.05115	10.014
#2	.40717	10.200	.39474	.49519	1.0073	.05125	10.006
#3	.40630	10.191	.39992	.49250	1.0159	.05094	10.196
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05085	.19930	.50343	.51248	4.0210	48.769	1.0170
Stddev	.00016	.00069	.00173	.00042	.0444	.141	.0095
%RSD	.31221	.34558	.34304	.08145	1.1049	.29004	.93218
#1	.05072	.20005	.50491	.51259	3.9923	48.842	1.0084
#2	.05079	.19915	.50386	.51282	3.9985	48.605	1.0154
#3	.05103	.19869	.50153	.51201	4.0722	48.858	1.0271
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.6470	.50568	1.0217	49.589	.50657	9.9463	.50897
Stddev	.0505	.00379	.0050	.266	.00174	.0145	.00224
%RSD	.52402	.74931	.49273	.53683	.34444	.14541	.44016
#1	9.7040	.50521	1.0275	49.574	.50855	9.9602	.50639
#2	9.6294	.50215	1.0196	49.330	.50589	9.9313	.51042
#3	9.6076	.50969	1.0181	49.862	.50527	9.9474	.51010
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 18:08:15 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1949**	**.39405**	**5.1402**	**1.0022**	**1.0107**	**.99866**	**.51518**
Stddev	.0031	.00179	.0100	.0005	.0053	.01659	.00058
%RSD	.25554	.45308	.19427	.05270	.52786	1.6608	.11222
#1	1.1981	.39492	5.1482	1.0028	1.0100	1.0024	.51475
#2	1.1944	.39525	5.1290	1.0020	1.0058	.98052	.51584
#3	1.1920	.39200	5.1434	1.0018	1.0164	1.0131	.51497
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0309**	**1.0070**	F **2.9451**
Stddev	.0008	.0014	.0223
%RSD	.08066	.13880	.75803
#1	1.0316	1.0084	2.9555
#2	1.0300	1.0069	2.9195
#3	1.0311	1.0056	2.9603
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9649.2**	**58096.**	**2982.6**
Stddev	5.7	114.	17.5
%RSD	.05915	.19560	.58802
#1	9648.4	58223.	2978.9
#2	9644.0	58003.	3001.7
#3	9655.3	58062.	2967.2

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 18:11:52 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00062**	**.00113**	**.00060**	**-.00134**	**.00269**	**.00016**	**-.00509**	**-.00011**
Stddev	.00447	.01287	.00139	.00141	.00058	.00006	.00999	.00008
%RSD	723.56	1142.0	234.30	105.23	21.456	39.000	196.19	75.399
#1	-.00262	.00580	.00212	.00028	.00233	.00022	-.01646	-.00020
#2	.00450	-.01343	-.00061	-.00230	.00238	.00017	.00228	-.00009
#3	-.00373	.01101	.00028	-.00199	.00335	.00009	-.00109	-.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00010**	**.00031**	**.00131**	**.01880**	**-.14052**	**-.00173**	**.03086**	**-.00243**
Stddev	.00043	.00144	.00103	.01556	.08453	.00290	.07623	.00189
%RSD	438.61	470.86	78.889	82.745	60.152	168.14	247.00	77.536
#1	.00038	.00014	.00051	.03667	-.04368	.00153	-.05463	-.00186
#2	-.00045	-.00104	.00248	.01148	-.17839	-.00266	.09175	-.00090
#3	-.00022	.00182	.00094	.00826	-.19949	-.00404	.05546	-.00454
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00375**	**-.00702**	**.00000**	**.00544**	**-.00002**	**.00180**	**.00158**	**.03715**
Stddev	.00019	.01848	.00030	.00537	.00129	.00320	.00255	.00232
%RSD	5.0585	263.40	8008.1	98.759	6600.5	177.29	161.59	6.2433
#1	.00382	-.02711	-.00030	.01149	.00142	.00145	-.00134	.03895
#2	.00354	-.00319	.00002	.00357	-.00108	.00516	.00266	.03453
#3	.00390	.00925	.00029	.00125	-.00040	-.00120	.00341	.03795
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 18:11:52 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00026	.00036	-.00017	-.00186	.00075	.00012	.00425
Stddev	.00040	.00044	.01020	.00315	.00028	.00007	.00095
%RSD	153.60	122.95	5879.7	169.59	36.989	59.021	22.345
#1	.00000	.00086	.00219	-.00527	.00044	.00006	.00508
#2	.00073	.00017	.00864	-.00126	.00081	.00010	.00321
#3	.00006	.00004	-.01135	.00095	.00098	.00019	.00446
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9372.1	56632.	2874.9
Stddev	24.8	50.	14.1
%RSD	.26420	.08878	.48885
#1	9350.8	56589.	2886.0
#2	9366.2	56619.	2859.1
#3	9399.3	56687.	2879.7

Approved: June 27, 2016

Sample Name: L1606098610 Acquired: 6/25/2016 18:15:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00168	.00613	.00026	.00622	.01341	.00005	115.37	.00029
Stddev	.00132	.00327	.00219	.00120	.00108	.00003	.56	.00010
%RSD	78.702	53.342	840.89	19.355	8.0675	55.551	.48696	35.866
#1	.00017	.00551	-.00222	.00662	.01310	.00009	115.81	.00019
#2	.00222	.00322	.00107	.00487	.01252	.00003	114.73	.00040
#3	.00265	.00967	.00193	.00717	.01462	.00004	115.55	.00029
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00004	.00098	.00166	.06083	.38326	.00831	27.726	.01242
Stddev	.00045	.00011	.00061	.01132	.08932	.00154	.200	.00158
%RSD	1035.7	11.515	36.634	18.614	23.306	18.467	.72137	12.708
#1	-.00034	.00088	.00210	.06356	.35599	.00912	27.699	.01068
#2	-.00027	.00110	.00191	.04839	.48304	.00927	27.542	.01377
#3	.00048	.00096	.00096	.07054	.31076	.00654	27.939	.01280
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00076	14.397	.00169	.00670	.00176	-.00418	.00480	3.9389
Stddev	.00003	.096	.00045	.00675	.00408	.00345	.00343	.0208
%RSD	3.3704	.66673	26.485	100.67	231.83	82.446	71.530	.52736
#1	.00075	14.373	.00205	.00909	-.00024	-.00395	.00094	3.9592
#2	.00074	14.315	.00119	.01193	.00646	-.00774	.00595	3.9398
#3	.00079	14.503	.00185	-.00091	-.00093	-.00086	.00752	3.9176
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098610 Acquired: 6/25/2016 18:15:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00034**	**.52737**	**-.01662**	**-.00476**	**.00053**	**.01128**	**.00653**
Stddev	.00061	.00238	.00515	.00157	.00069	.00008	.00193
%RSD	181.06	.45087	31.015	32.913	130.43	.73884	29.481
#1	-.00078	.52702	-.01240	-.00297	.00012	.01121	.00451
#2	.00036	.52519	-.01509	-.00547	.00132	.01137	.00674
#3	-.00059	.52991	-.02236	-.00585	.00014	.01126	.00835
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9795.7**	**59337.**	**3075.9**
Stddev	21.1	178.	25.1
%RSD	.21493	.29915	.81568
#1	9771.4	59262.	3048.2
#2	9807.4	59210.	3082.7
#3	9808.3	59540.	3097.0

Approved: June 27, 2016

Sample Name: L1606098612 Acquired: 6/25/2016 18:19:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00237	.00267	-.00060	.08716	.06662	.00041	142.66	.00007
Stddev	.00073	.00275	.00337	.00141	.00149	.00002	1.19	.00021
%RSD	30.688	102.87	562.71	1.6192	2.2393	3.9782	.83233	319.26
#1	.00315	.00518	-.00334	.08576	.06575	.00041	141.33	.00027
#2	.00225	-.00027	-.00161	.08858	.06577	.00039	143.01	.00007
#3	.00171	.00310	.00316	.08713	.06835	.00042	143.63	-.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00010	.00135	.04528	.02649	1.6602	.01963	40.720	.03253
Stddev	.00034	.00053	.00134	.03733	.1018	.00353	.262	.00324
%RSD	339.50	39.036	2.9680	140.93	6.1325	17.994	.64244	9.9559
#1	.00034	.00170	.04668	.04343	1.6128	.02218	40.440	.02879
#2	-.00029	.00162	.04400	-.01631	1.5907	.02112	40.763	.03448
#3	.00024	.00075	.04516	.05234	1.7770	.01560	40.958	.03432
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00115	26.540	.00106	.02312	.00172	-.00084	.00204	2.4275
Stddev	.00029	.172	.00094	.00593	.00284	.00147	.00269	.0046
%RSD	25.080	.64693	88.979	25.661	164.80	174.96	131.78	.18880
#1	.00089	26.345	.00153	.02349	.00164	-.00125	.00490	2.4289
#2	.00110	26.607	.00167	.01701	.00461	-.00205	-.00043	2.4311
#3	.00146	26.668	-.00003	.02886	-.00107	.00079	.00165	2.4223
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098612 Acquired: 6/25/2016 18:19:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00022**	**1.7410**	**-.01626**	**-.00241**	**.00038**	**.05602**	**.00415**
Stddev	.00061	.0138	.00313	.00137	.00051	.00009	.00053
%RSD	273.86	.79166	19.237	56.738	134.14	.16489	12.835
#1	-.00019	1.7255	-.01265	-.00306	.00050	.05591	.00443
#2	.00092	1.7457	-.01795	-.00084	.00081	.05608	.00354
#3	-.00006	1.7519	-.01818	-.00334	-.00018	.05607	.00449
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9747.5**	**58989.**	**3047.5**
Stddev	4.1	78.	13.8
%RSD	.04198	.13200	.45213
#1	9749.1	58936.	3063.4
#2	9750.6	59079.	3039.1
#3	9742.9	58952.	3040.0

Approved: June 27, 2016

Sample Name: L1606098614 Acquired: 6/25/2016 18:23:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00279	.01008	-.00066	.03390	.03124	.00010	196.20	.00021
Stddev	.00112	.00383	.00147	.00206	.00013	.00007	.85	.00008
%RSD	40.043	38.038	221.13	6.0906	.42986	73.782	.43251	39.178
#1	.00183	.00766	.00090	.03506	.03108	.00016	195.71	.00030
#2	.00402	.00807	-.00201	.03513	.03132	.00002	195.70	.00014
#3	.00253	.01449	-.00089	.03152	.03131	.00013	197.18	.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	.00088	.00840	.02428	2.9211	.02334	75.647	.04008
Stddev	.00023	.00045	.00064	.00605	.0291	.00193	.677	.00318
%RSD	446.71	50.770	7.6701	24.929	.99693	8.2676	.89445	7.9254
#1	-.00007	.00063	.00780	.03006	2.8958	.02251	74.950	.04023
#2	.00031	.00062	.00908	.02477	2.9529	.02197	75.691	.03683
#3	-.00009	.00140	.00832	.01799	2.9145	.02555	76.301	.04318
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00071	22.556	.00040	.01184	.00385	-.00016	.00418	3.4833
Stddev	.00033	.097	.00156	.00464	.00193	.00051	.00681	.0215
%RSD	46.348	.43032	390.32	39.236	50.052	315.18	162.82	.61768
#1	.00068	22.526	-.00110	.00655	.00167	-.00006	.01171	3.5066
#2	.00040	22.478	.00029	.01372	.00533	-.00072	-.00154	3.4792
#3	.00106	22.665	.00201	.01524	.00455	.00029	.00237	3.4641
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098614 Acquired: 6/25/2016 18:23:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**2.3670**	**-.02425**	**-.00232**	**-.00024**	**.02160**	**.00232**
Stddev	.00092	.0076	.00398	.00117	.00100	.00003	.00161
%RSD	715.81	.32195	16.401	50.456	415.47	.12320	69.325
#1	.00091	2.3659	-.02580	-.00236	.00091	.02157	.00413
#2	-.00082	2.3599	-.02722	-.00113	-.00075	.02159	.00107
#3	-.00048	2.3751	-.01973	-.00348	-.00089	.02163	.00175
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9640.5**	**58425.**	**3058.6**
Stddev	20.6	100.	4.7
%RSD	.21337	.17110	.15390
#1	9631.4	58473.	3059.7
#2	9664.1	58492.	3062.6
#3	9626.1	58310.	3053.4

Approved: June 27, 2016

Sample Name: L1606098616 Acquired: 6/25/2016 18:27:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00115	.03265	.00256	.11861	.01251	.00029	229.75	.00020
Stddev	.00195	.00899	.00533	.00161	.00061	.00014	.36	.00016
%RSD	169.27	27.532	208.27	1.3580	4.8819	47.504	.15530	78.536
#1	.00069	.04218	-.00197	.11885	.01280	.00014	229.65	.00038
#2	-.00053	.02433	.00844	.12008	.01292	.00030	229.45	.00006
#3	.00329	.03143	.00122	.11689	.01181	.00042	230.15	.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00062	.00178	.06744	.03108	3.2126	.03682	84.360	.29574
Stddev	.00023	.00069	.00047	.01537	.1307	.00386	.426	.00133
%RSD	37.410	38.666	.69994	49.454	4.0677	10.480	.50488	.44809
#1	.00042	.00192	.06781	.02007	3.0741	.03526	83.896	.29633
#2	.00087	.00239	.06691	.04863	3.2299	.03400	84.450	.29422
#3	.00056	.00104	.06760	.02452	3.3338	.04122	84.734	.29667
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00072	71.143	.00123	.03078	.00375	-.00200	.00018	5.1619
Stddev	.00027	.049	.00055	.00199	.00114	.00238	.00481	.0107
%RSD	37.642	.06818	44.575	6.4658	30.424	119.08	2715.0	.20775
#1	.00061	71.192	.00158	.03108	.00244	-.00009	.00517	5.1511
#2	.00102	71.095	.00060	.02865	.00425	-.00467	-.00442	5.1621
#3	.00051	71.141	.00151	.03260	.00456	-.00125	-.00023	5.1725
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098616 Acquired: 6/25/2016 18:27:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00037	6.5212	-.02963	-.00177	-.00109	.08946	.00332
Stddev	.00060	.0079	.00654	.00230	.00079	.00019	.00190
%RSD	161.30	.12151	22.085	130.49	72.258	.21177	57.202
#1	.00105	6.5204	-.02657	.00051	-.00123	.08936	.00540
#2	.00018	6.5138	-.02518	-.00171	-.00179	.08968	.00287
#3	-.00011	6.5296	-.03715	-.00409	-.00024	.08933	.00168
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9566.8	58078.	3050.1
Stddev	16.9	52.	19.7
%RSD	.17631	.08955	.64584
#1	9563.8	58018.	3068.4
#2	9585.0	58111.	3029.2
#3	9551.7	58106.	3052.7

Approved: June 27, 2016

Sample Name: L1606098618 Acquired: 6/25/2016 18:31:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00155	-.00096	.00289	.04251	.02915	.00005	62.293	-.00011
Stddev	.00038	.00238	.00029	.00109	.00082	.00008	.092	.00019
%RSD	24.410	246.77	9.9664	2.5737	2.7985	149.80	.14819	166.73
#1	.00113	.00168	.00257	.04189	.03009	-.00003	62.335	-.00017
#2	.00185	-.00164	.00314	.04186	.02874	.00005	62.357	.00010
#3	.00168	-.00294	.00296	.04377	.02862	.00013	62.187	-.00027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.00031	.00347	.01624	1.9071	.02198	20.259	.01093
Stddev	.00021	.00046	.00016	.00938	.1214	.00171	.067	.00376
%RSD	243.30	146.24	4.6588	57.772	6.3644	7.7782	.33311	34.347
#1	.00028	.00083	.00329	.00573	1.8306	.02394	20.336	.01303
#2	.00011	.00015	.00350	.01922	2.0470	.02118	20.230	.00660
#3	-.00013	-.00004	.00361	.02377	1.8435	.02082	20.211	.01318
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00102	25.999	-.00042	.00444	.00244	-.00359	.00392	3.1846
Stddev	.00023	.100	.00069	.00545	.00095	.00188	.00301	.0187
%RSD	22.864	.38558	161.73	122.87	39.121	52.433	76.809	.58628
#1	.00106	26.085	-.00101	-.00067	.00273	-.00311	.00526	3.2018
#2	.00123	26.024	.00033	.00381	.00321	-.00199	.00047	3.1874
#3	.00077	25.889	-.00059	.01018	.00137	-.00566	.00602	3.1647
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098618 Acquired: 6/25/2016 18:31:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00057	1.4545	-.00154	-.00233	.00037	.00832	.00208
Stddev	.00132	.0032	.00357	.00066	.00141	.00007	.00021
%RSD	228.95	.22347	232.21	28.221	386.86	.88012	10.206
#1	.00206	1.4546	-.00325	-.00241	-.00025	.00825	.00199
#2	-.00043	1.4577	-.00393	-.00294	-.00063	.00831	.00233
#3	.00009	1.4512	.00257	-.00163	.00198	.00839	.00193
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9841.2	59299.	3061.4
Stddev	105.5	272.	4.9
%RSD	1.0718	.45797	.16152
#1	9746.5	58986.	3061.4
#2	9822.2	59458.	3056.4
#3	9954.9	59454.	3066.3

Approved: June 27, 2016

Sample Name: L1606098620 Acquired: 6/25/2016 18:35:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00004	.03312	.00066	.07608	.03474	.00017	4.5650	.00010
Stddev	.00141	.00304	.00168	.00356	.00133	.00010	.0156	.00016
%RSD	3777.7	9.1736	255.45	4.6823	3.8419	59.446	.34129	166.50
#1	-.00004	.03663	-.00075	.08018	.03521	.00029	4.5712	.00011
#2	.00149	.03142	.00252	.07372	.03577	.00014	4.5766	.00025
#3	-.00133	.03132	.00020	.07433	.03323	.00009	4.5473	-.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00112	.00060	.00299	.05475	.56414	.00988	1.7714	.01450
Stddev	.00021	.00082	.00143	.01752	.07965	.00243	.0462	.00322
%RSD	18.384	137.46	47.864	31.992	14.119	24.596	2.6071	22.223
#1	.00132	.00110	.00244	.06300	.59505	.00894	1.8247	.01687
#2	.00115	.00105	.00462	.03463	.62370	.00806	1.7436	.01579
#3	.00091	-.00035	.00192	.06662	.47367	.01264	1.7459	.01083
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00137	251.01	-.00057	.03777	.00214	-.00589	.00108	2.0603
Stddev	.00029	.75	.00178	.00427	.00057	.00380	.00694	.0083
%RSD	21.156	.29997	313.28	11.296	26.364	64.542	640.75	.40056
#1	.00125	251.39	-.00025	.04217	.00279	-.00381	-.00651	2.0621
#2	.00170	251.50	-.00248	.03365	.00178	-.01028	.00710	2.0676
#3	.00115	250.15	.00103	.03749	.00186	-.00358	.00266	2.0514
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098620 Acquired: 6/25/2016 18:35:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00101**	**.14703**	**.00470**	**-.00231**	**.00100**	**.00923**	**.00423**
Stddev	.00046	.00057	.00299	.00179	.00041	.00019	.00094
%RSD	44.993	.38637	63.686	77.674	41.001	2.0938	22.219
#1	.00049	.14737	.00735	-.00044	.00130	.00917	.00391
#2	.00129	.14734	.00531	-.00246	.00116	.00906	.00349
#3	.00126	.14637	.00145	-.00402	.00053	.00944	.00529
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9725.3**	**58367.**	**3028.8**
Stddev	17.5	120.	15.3
%RSD	.18040	.20576	.50451
#1	9742.7	58420.	3020.7
#2	9707.6	58451.	3019.2
#3	9725.6	58230.	3046.4

Approved: June 27, 2016

Sample Name: L1606098622 Acquired: 6/25/2016 18:39:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00117	-.00596	.00037	.05747	.02764	.00023	127.72	.00011
Stddev	.00093	.00925	.00122	.00204	.00088	.00005	1.04	.00012
%RSD	79.226	155.22	327.64	3.5486	3.1759	23.660	.81279	108.09
#1	.00125	-.01464	.00026	.05716	.02813	.00022	127.35	.00021
#2	.00205	-.00701	-.00078	.05560	.02662	.00018	126.92	.00013
#3	.00020	.00377	.00164	.05965	.02815	.00028	128.89	-.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00027	.00090	.01009	.02031	2.5853	.01995	32.558	.04793
Stddev	.00056	.00109	.00042	.00894	.0945	.00518	.278	.00190
%RSD	208.67	121.92	4.1494	44.026	3.6562	25.950	.85453	3.9665
#1	.00085	.00146	.01042	.02811	2.6682	.02326	32.522	.05007
#2	.00021	-.00036	.01022	.02229	2.4824	.01399	32.299	.04730
#3	-.00026	.00160	.00962	.01055	2.6053	.02261	32.852	.04643
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00038	46.975	.00038	.00992	.00378	-.00381	.00198	3.5382
Stddev	.00001	.389	.00033	.00514	.00270	.00385	.00283	.0119
%RSD	1.5146	.82774	87.053	51.791	71.542	100.88	142.89	.33659
#1	.00038	46.842	.00000	.00464	.00569	-.00293	.00001	3.5245
#2	.00039	46.670	.00061	.01491	.00068	-.00048	.00522	3.5450
#3	.00039	47.413	.00052	.01022	.00497	-.00802	.00070	3.5451
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098622 Acquired: 6/25/2016 18:39:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00017	1.7827	-.01653	-.00317	-.00057	.00858	.00262
Stddev	.00056	.0144	.00634	.00502	.00088	.00017	.00024
%RSD	326.45	.80805	38.350	158.35	154.19	1.9458	9.0067
#1	-.00048	1.7747	-.02093	-.00538	-.00155	.00859	.00236
#2	.00051	1.7740	-.00927	.00258	.00014	.00841	.00281
#3	.00049	1.7993	-.01940	-.00671	-.00029	.00875	.00270
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9602.2	57749.	2959.9
Stddev	40.4	252.	26.3
%RSD	.42032	.43708	.88902
#1	9555.9	57561.	2977.4
#2	9630.3	57650.	2972.6
#3	9620.3	58036.	2929.6

Approved: June 27, 2016

Sample Name: L1606098624 Acquired: 6/25/2016 18:42:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00037**	**.00561**	**.00154**	**.40348**	**.02348**	**.00039**	**237.24**
Stddev	.00142	.00394	.00506	.00070	.00039	.00006	.43
%RSD	386.92	70.246	327.66	.17357	1.6661	16.452	.18001
#1	.00014	.00128	.00004	.40369	.02381	.00037	237.56
#2	-.00093	.00658	-.00259	.40270	.02305	.00046	237.41
#3	.00189	.00898	.00718	.40405	.02358	.00034	236.75
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**-.00007**	**.00152**	**.02649**	**.02448**	**5.9223**	**.01174**
Stddev	.00017	.00052	.00056	.00200	.00699	.0587	.00396
%RSD	362.68	725.31	36.507	7.5377	28.540	.99123	33.722
#1	-.00014	.00052	.00185	.02553	.03123	5.8563	.01398
#2	-.00015	-.00040	.00183	.02879	.02493	5.9416	.00717
#3	.00015	-.00034	.00088	.02516	.01728	5.9689	.01408
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**66.698**	**.32515**	**.00019**	**63.762**	**.00021**	**.01502**	**.00144**
Stddev	.426	.00095	.00039	.087	.00030	.00323	.00065
%RSD	.63897	.29073	207.41	.13653	145.82	21.500	44.732
#1	66.222	.32513	-.00023	63.794	-.00012	.01765	.00108
#2	67.045	.32421	.00055	63.829	.00049	.01601	.00219
#3	66.826	.32610	.00024	63.663	.00025	.01142	.00106
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: L1606098624 Acquired: 6/25/2016 18:42:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-02

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00149**	**.00522**	**3.3709**	**-.00072**	**1.6690**	F **-.03355**	**-.00155**
Stddev	.00099	.00177	.0249	.00122	.0015	.00264	.00229
%RSD	66.177	33.876	.73770	169.83	.08806	7.8818	147.53
#1	-.00080	.00419	3.3514	-.00109	1.6686	-.03207	.00107
#2	-.00262	.00726	3.3989	-.00171	1.6706	-.03199	-.00316
#3	-.00106	.00421	3.3624	.00064	1.6678	-.03661	-.00256
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00040**	**.00877**	**.00210**
Stddev	.00049	.00015	.00054
%RSD	123.14	1.7509	25.986
#1	-.00082	.00894	.00208
#2	.00014	.00865	.00156
#3	-.00050	.00870	.00265
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9491.7**	**57630.**	**3009.2**
Stddev	78.4	156.	15.6
%RSD	.82592	.27138	.51996
#1	9550.0	57451.	2991.3
#2	9402.6	57694.	3015.7
#3	9522.6	57743.	3020.5

Approved: June 27, 2016

Sample Name: L1606098626 Acquired: 6/25/2016 18:46:49 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00327	.00679	.00038	.03972	.02631	.00014	117.32	.00023
Stddev	.00072	.00538	.00107	.00258	.00068	.00005	.34	.00019
%RSD	22.022	79.326	284.61	6.5079	2.6006	35.377	.28683	83.944
#1	.00409	.00166	-.00067	.03695	.02674	.00014	117.70	.00003
#2	.00295	.01240	.00033	.04014	.02552	.00019	117.19	.00023
#3	.00276	.00630	.00147	.04206	.02666	.00009	117.07	.00042
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00020	.00126	.00604	.03025	1.5600	.01598	35.418	-.00162
Stddev	.00014	.00104	.00157	.00896	.0430	.00117	.155	.00177
%RSD	71.567	82.474	25.938	29.612	2.7589	7.3145	.43775	108.92
#1	.00004	.00238	.00775	.03493	1.5696	.01501	35.426	-.00126
#2	.00029	.00033	.00573	.01992	1.5974	.01567	35.569	-.00006
#3	.00026	.00107	.00466	.03588	1.5130	.01728	35.259	-.00354
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00045	17.994	-.00024	.00950	.00273	-.00253	.00429	2.8578
Stddev	.00029	.066	.00113	.00146	.00267	.00313	.00183	.0046
%RSD	65.363	.36928	461.23	15.318	98.010	123.57	42.598	.16060
#1	.00011	18.060	.00031	.00854	.00072	.00062	.00280	2.8530
#2	.00057	17.927	.00050	.00879	.00170	-.00564	.00374	2.8621
#3	.00066	17.996	-.00154	.01118	.00576	-.00258	.00633	2.8583
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098626 Acquired: 6/25/2016 18:46:49 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00012	.95352	-.01676	-.00181	.00025	.00656	.00282
Stddev	.00004	.00361	.00400	.00055	.00086	.00005	.00124
%RSD	36.480	.37812	23.874	30.252	341.08	.79230	44.156
#1	.00017	.95765	-.01632	-.00120	.00116	.00654	.00348
#2	.00013	.95102	-.02097	-.00198	-.00055	.00661	.00138
#3	.00008	.95189	-.01300	-.00225	.00014	.00651	.00358
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9709.6	59301.	3058.5
Stddev	45.7	137.	18.4
%RSD	.47038	.23115	.60138
#1	9758.2	59149.	3039.9
#2	9703.1	59336.	3058.7
#3	9667.6	59417.	3076.7

Approved: June 27, 2016

Sample Name: L1606098628 Acquired: 6/25/2016 18:50:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00127**	**.00062**	**.00226**	**.03408**	**.02631**	**.00011**	**202.60**
Stddev	.00142	.00505	.00320	.00146	.00145	.00006	1.09
%RSD	111.97	810.26	141.74	4.2925	5.4989	52.470	.53642
#1	.00192	-.00172	.00570	.03364	.02795	.00015	202.20
#2	.00225	-.00283	-.00063	.03289	.02521	.00012	201.76
#3	-.00036	.00642	.00171	.03571	.02576	.00005	203.83
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**-.00018**	**.00230**	**.00706**	**.02884**	**3.1108**	**.02231**
Stddev	.00015	.00008	.00109	.00066	.02764	.1253	.00468
%RSD	173.92	43.590	47.288	9.3696	95.862	4.0286	20.962
#1	.00015	-.00019	.00355	.00741	.06027	3.2199	.01700
#2	-.00008	-.00010	.00186	.00630	.00830	2.9739	.02412
#3	.00019	-.00026	.00151	.00747	.01794	3.1385	.02581
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**77.982**	**.03186**	**.00054**	**22.893**	**.00124**	**.00646**	**.00417**
Stddev	.421	.00093	.00029	.135	.00089	.00201	.00077
%RSD	.53979	2.9039	53.423	.59091	71.804	31.048	18.563
#1	77.763	.03137	.00064	22.869	.00109	.00439	.00500
#2	77.715	.03128	.00078	22.771	.00220	.00660	.00405
#3	78.467	.03292	.00022	23.039	.00044	.00839	.00347
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: L1606098628 Acquired: 6/25/2016 18:50:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00202**	**.00107**	**3.6233**	**.00060**	**2.4441**	F **-.03051**	**-.00323**
Stddev	.00386	.00287	.0015	.00175	.0120	.00406	.00089
%RSD	190.96	268.98	.04080	291.82	.49257	13.319	27.545
#1	.00200	.00365	3.6232	-.00046	2.4407	-.03513	-.00322
#2	-.00236	-.00203	3.6248	-.00036	2.4341	-.02890	-.00413
#3	-.00571	.00158	3.6218	.00262	2.4575	-.02750	-.00235
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00057**	**.01373**	**.00279**
Stddev	.00055	.00005	.00121
%RSD	96.088	.39414	43.503
#1	-.00086	.01376	.00139
#2	-.00092	.01376	.00355
#3	.00006	.01367	.00343
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9549.4**	**58061.**	**2994.7**
Stddev	10.8	342.	16.3
%RSD	.11322	.58988	.54589
#1	9556.3	57786.	2984.2
#2	9554.9	57953.	3013.5
#3	9536.9	58445.	2986.3

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 18:54:37 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.36142	9.0966	F .35164	F .43504	F .88841	.04556	F 8.8985
Stddev	.00063	.0241	.00214	.00359	.00668	.00015	.0636
%RSD	.17458	.26527	.60772	.82605	.75211	.32062	.71530
#1	.36081	9.1069	.35253	.43880	.88074	.04567	8.8250
#2	.36138	9.0691	.35318	.43467	.89150	.04539	8.9365
#3	.36207	9.1139	.34920	.43164	.89299	.04562	8.9341
Check ?	Chk Pass	Chk Pass	Chk Fail	Chk Fail	Chk Fail	Chk Pass	Chk Fail
Value			.40000	.50000	1.0000		10.000
Range			-10.000%	-10.000%	-10.000%		-10.000%

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04504	F .17683	F .44843	.45584	F 3.5682	F 43.223	.90012
Stddev	.00036	.00059	.00210	.00028	.0180	.402	.00642
%RSD	.79578	.33417	.46848	.06078	.50416	.93081	.71346
#1	.04493	.17615	.45006	.45571	3.5475	42.758	.89283
#2	.04475	.17711	.44916	.45565	3.5796	43.446	.90262
#3	.04544	.17723	.44606	.45616	3.5775	43.464	.90492
Check ?	Chk Pass	Chk Fail	Chk Fail	Chk Pass	Chk Fail	Chk Fail	Chk Pass
Value		.20000	.50000		4.0000	50.000	
Range		-10.000%	-10.000%		-10.000%	-10.000%	

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F 8.4568	F .44683	.90610	F 43.837	.45006	F 8.8106	F .44997
Stddev	.1941	.00153	.00307	.290	.00031	.0184	.00285
%RSD	2.2957	.34349	.33891	.66256	.06941	.20845	.63417
#1	8.2816	.44509	.90942	43.504	.44975	8.7897	.44790
#2	8.6655	.44738	.90552	43.973	.45037	8.8181	.44879
#3	8.4231	.44801	.90336	44.035	.45007	8.8241	.45322
Check ?	Chk Fail	Chk Fail	Chk Pass	Chk Fail	Chk Pass	Chk Fail	Chk Fail
Value	10.000	.50000		50.000		10.000	.50000
Range	-10.000%	-10.000%		-10.000%		-10.000%	-10.000%

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 18:54:37 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F 1.0511	F .35868	4.5430	F .88663	F .88982	F .89732	.46317
Stddev	.0038	.00393	.0064	.00114	.00700	.00126	.00115
%RSD	.35733	1.0963	.14052	.12909	.78679	.14039	.24883
#1	1.0514	.35522	4.5358	.88537	.88186	.89705	.46442
#2	1.0547	.35787	4.5453	.88691	.89256	.89622	.46295
#3	1.0472	.36295	4.5480	.88760	.89503	.89870	.46215
Check ?	Chk Fail	Chk Fail	Chk Pass	Chk Fail	Chk Fail	Chk Fail	Chk Pass
Value	1.2000	.40000		1.0000	1.0000	1.0000	
Range	-10.000%	-10.000%		-10.000%	-10.000%	-10.000%	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.92451	F .89326	F 2.6142
Stddev	.00239	.00079	.0303
%RSD	.25806	.08826	1.1591
#1	.92697	.89416	2.5794
#2	.92435	.89291	2.6285
#3	.92221	.89270	2.6347
Check ?	Chk Pass	Chk Fail	Chk Fail
Value		1.0000	1.0000
Range		-10.000%	10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10436.	62603.	3229.6
Stddev	7.	58.	39.3
%RSD	.06851	.09259	1.2179
#1	10435.	62597.	3274.8
#2	10443.	62548.	3203.4
#3	10429.	62663.	3210.5

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 18:58:15 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00102	-.00634	.00315	-.00143	.00123	.00009	-.03120	-.00001
Stddev	.00036	.00615	.00146	.00223	.00014	.00007	.03757	.00022
%RSD	35.623	96.910	46.271	156.16	11.328	79.889	120.40	3000.5
#1	.00142	-.00623	.00470	-.00112	.00129	.00016	.01024	.00017
#2	.00091	-.01255	.00293	-.00379	.00133	.00004	-.06303	-.00025
#3	.00072	-.00026	.00181	.00063	.00107	.00005	-.04081	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00001	-.00012	.00040	.00652	.00729	-.00094	.01960	-.00009
Stddev	.00003	.00070	.00057	.00448	.10732	.00258	.12813	.00039
%RSD	340.66	581.54	142.03	68.614	1471.3	275.05	653.72	408.22
#1	-.00000	.00019	-.00025	.00247	.06763	-.00204	-.01079	.00011
#2	.00004	-.00093	.00068	.01133	.07087	.00201	.16019	-.00054
#3	-.00001	.00037	.00077	.00577	-.11662	-.00279	-.09060	.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00312	.01765	-.00021	.00497	.00440	.00279	.00542	.03094
Stddev	.00023	.01443	.00123	.00520	.00590	.00458	.00181	.00198
%RSD	7.4573	81.775	586.87	104.64	134.12	164.15	33.357	6.4039
#1	.00298	.03423	-.00101	.00374	.00273	.00293	.00481	.03016
#2	.00339	.00793	-.00083	.01067	.01095	-.00186	.00399	.03319
#3	.00300	.01078	.00121	.00049	-.00049	.00730	.00745	.02947
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 18:58:15 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00029**	**.00057**	**.00328**	**-.00363**	**.00052**	**.00010**	**.00222**
Stddev	.00020	.00050	.00575	.00205	.00106	.00011	.00064
%RSD	66.832	88.088	175.09	56.396	206.25	116.18	28.891
#1	.00017	.00114	-.00137	-.00544	-.00047	.00019	.00164
#2	.00052	.00028	.00971	-.00141	.00038	-.00003	.00291
#3	.00019	.00027	.00151	-.00405	.00164	.00012	.00210
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10029.**	**60890.**	**3064.8**
Stddev	14.	97.	13.9
%RSD	.13680	.15850	.45397
#1	10045.	60780.	3055.3
#2	10022.	60930.	3080.8
#3	10020.	60960.	3058.3

Approved: June 27, 2016

Sample Name: L1606098630 Acquired: 6/25/2016 19:02:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00160**	**.04734**	**-.00156**	**.13569**	**.01871**	**.00015**	**238.04**
Stddev	.00181	.00348	.00404	.00135	.00123	.00008	.85
%RSD	113.48	7.3553	258.52	.99665	6.5967	52.824	.35679
#1	.00017	.04905	-.00155	.13421	.01822	.00022	239.01
#2	.00364	.04964	.00247	.13685	.01779	.00007	237.71
#3	.00098	.04333	-.00561	.13603	.02011	.00014	237.41
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00026**	**.00032**	**.00251**	**.01182**	**.03060**	**3.4183**	**.04090**
Stddev	.00010	.00018	.00081	.00072	.01332	.0462	.00200
%RSD	39.577	56.009	32.204	6.0677	43.540	1.3506	4.8796
#1	.00014	.00047	.00287	.01260	.01963	3.3856	.03874
#2	.00034	.00036	.00308	.01119	.04543	3.4711	.04128
#3	.00028	.00012	.00159	.01167	.02674	3.3981	.04268
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**90.439**	**.21655**	**.00156**	**78.879**	**.00140**	**.01756**	**.00033**
Stddev	.173	.00121	.00013	.416	.00068	.00677	.00046
%RSD	.19086	.55720	8.0204	.52709	48.349	38.536	138.30
#1	90.626	.21726	.00143	79.353	.00191	.01647	.00076
#2	90.405	.21515	.00168	78.710	.00063	.02480	-.00016
#3	90.286	.21722	.00157	78.574	.00166	.01140	.00040
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: L1606098630 Acquired: 6/25/2016 19:02:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00294**	**-.00177**	**5.6111**	**-.00017**	**7.0275**	F **-.03316**	**-.00254**
Stddev	.00229	.00087	.0076	.00038	.0379	.00293	.00099
%RSD	77.743	49.294	.13521	227.38	.53946	8.8247	38.798
#1	-.00354	-.00250	5.6121	-.00026	7.0707	-.03043	-.00171
#2	-.00042	-.00199	5.6030	-.00050	7.0124	-.03625	-.00363
#3	-.00488	-.00081	5.6181	.00025	6.9995	-.03280	-.00229
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00065**	**.02003**	**.00667**
Stddev	.00064	.00007	.00116
%RSD	98.374	.34297	17.407
#1	-.00116	.01996	.00557
#2	-.00088	.02003	.00654
#3	.00007	.02010	.00788
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9361.9**	**56594.**	**2990.6**
Stddev	23.4	147.	20.1
%RSD	.24995	.25968	.67365
#1	9379.7	56762.	2967.6
#2	9370.7	56528.	3005.2
#3	9335.4	56491.	2998.9

Approved: June 27, 2016

Sample Name: L1606098632S Acquired: 6/25/2016 19:06:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-06

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.18902**	**4.8212**	**.19116**	**.97374**	**.50077**	**.02402**	**134.52**	**.02388**
Stddev	.00088	.0081	.00239	.00462	.00118	.00003	.34	.00017
%RSD	.46488	.16867	1.2481	.47438	.23617	.14314	.25228	.70472
#1	.18980	4.8271	.19377	.97198	.50123	.02406	134.68	.02373
#2	.18919	4.8246	.18910	.97898	.49943	.02400	134.13	.02406
#3	.18807	4.8119	.19060	.97026	.50165	.02400	134.75	.02386
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09189**	**.23665**	**.24218**	**1.8859**	**25.617**	**.51395**	**37.596**	**.28166**
Stddev	.00021	.00120	.00028	.0089	.037	.00043	.192	.00203
%RSD	.23170	.50673	.11712	.47202	.14467	.08375	.50984	.71945
#1	.09190	.23534	.24190	1.8828	25.580	.51444	37.585	.28070
#2	.09167	.23691	.24247	1.8960	25.619	.51376	37.792	.28399
#3	.09210	.23769	.24219	1.8790	25.654	.51365	37.409	.28029
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49448**	**71.230**	**.23175**	**4.7310**	**.23416**	**.56777**	**.18400**	**6.1097**
Stddev	.00088	.164	.00128	.0060	.00191	.00602	.00319	.0038
%RSD	.17884	.22978	.55101	.12688	.81554	1.0604	1.7354	.06148
#1	.49529	71.252	.23067	4.7327	.23616	.57278	.18033	6.1129
#2	.49354	71.056	.23316	4.7244	.23394	.56944	.18613	6.1108
#3	.49461	71.382	.23142	4.7360	.23236	.56109	.18555	6.1056
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098632S Acquired: 6/25/2016 19:06:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-06

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.46869**	**2.2839**	**.45630**	**.23531**	**.49366**	**.46774**	**.01477**
Stddev	.00224	.0055	.00571	.00163	.00079	.00046	.00310
%RSD	.47718	.23955	1.2515	.69229	.16011	.09871	21.022
#1	.47047	2.2847	.45118	.23550	.49390	.46819	.01151
#2	.46942	2.2781	.45528	.23360	.49431	.46777	.01510
#3	.46618	2.2889	.46246	.23684	.49278	.46726	.01769
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9505.1**	**57443.**	**2995.1**
Stddev	12.6	181.	7.9
%RSD	.13263	.31567	.26290
#1	9490.9	57238.	2989.7
#2	9509.8	57583.	3004.2
#3	9514.8	57507.	2991.6

Approved: June 27, 2016

Sample Name: L1606098634S Acquired: 6/25/2016 19:09:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19165**	**4.8639**	**.19259**	**.98224**	**.50433**	**.02434**	**136.41**	**.02397**
Stddev	.00283	.0058	.00215	.00673	.00181	.00002	.41	.00039
%RSD	1.4742	.11932	1.1143	.68564	.35885	.08391	.29851	1.6453
#1	.19474	4.8576	.19410	.97512	.50399	.02434	136.06	.02438
#2	.19099	4.8690	.19014	.98851	.50272	.02431	136.30	.02394
#3	.18921	4.8651	.19354	.98310	.50629	.02435	136.86	.02359
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09370**	**.23944**	**.24457**	**1.9014**	**25.851**	**.51838**	**38.336**	**.27292**
Stddev	.00027	.00087	.00143	.0128	.044	.00156	.167	.00356
%RSD	.28910	.36336	.58497	.67456	.17178	.30011	.43499	1.3028
#1	.09395	.23880	.24603	1.9045	25.902	.51991	38.172	.27676
#2	.09341	.23909	.24449	1.9124	25.822	.51842	38.330	.27226
#3	.09373	.24043	.24318	1.8873	25.829	.51680	38.505	.26974
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49947**	**74.848**	**.23412**	**4.7760**	**.23373**	**.57493**	**.18272**	**6.1711**
Stddev	.00101	.224	.00114	.0149	.00097	.00427	.00635	.0140
%RSD	.20163	.29965	.48754	.31111	.41349	.74220	3.4752	.22725
#1	.50053	74.661	.23450	4.7820	.23289	.57819	.18571	6.1610
#2	.49852	74.786	.23283	4.7591	.23352	.57010	.18702	6.1652
#3	.49936	75.096	.23502	4.7869	.23478	.57650	.17542	6.1871
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098634S Acquired: 6/25/2016 19:09:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-07

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47482**	**2.3307**	**.45927**	**.23900**	**.49848**	**.47191**	**.00533**
Stddev	.00086	.0091	.01368	.00122	.00164	.00108	.00076
%RSD	.18068	.39132	2.9792	.51192	.32871	.22881	14.237
#1	.47581	2.3242	.47279	.23871	.49842	.47284	.00492
#2	.47440	2.3268	.44543	.23794	.50015	.47073	.00620
#3	.47425	2.3411	.45961	.24034	.49687	.47217	.00485
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9464.2**	**57182.**	**2975.2**
Stddev	11.1	55.	14.2
%RSD	.11755	.09672	.47580
#1	9463.1	57128.	2991.6
#2	9475.9	57180.	2967.1
#3	9453.7	57239.	2967.1

Approved: June 27, 2016

Sample Name: L1606098636SD Acquired: 6/25/2016 19:13:21 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19162**	**4.8756**	**.19532**	**1.3252**	**.50705**	**.02431**	**243.19**	**.02436**
Stddev	.00178	.0135	.00259	.0039	.00540	.00016	2.22	.00014
%RSD	.92905	.27765	1.3247	.29569	1.0646	.65896	.91281	.56637
#1	.19240	4.8805	.19500	1.3207	.51326	.02435	245.71	.02449
#2	.18958	4.8859	.19805	1.3282	.50441	.02444	242.33	.02421
#3	.19287	4.8603	.19291	1.3265	.50349	.02413	241.52	.02439
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09290**	**.23671**	**.29685**	**1.9145**	**29.596**	**.51587**	**71.839**	**.60737**
Stddev	.00022	.00172	.00059	.0369	.191	.00194	.252	.00417
%RSD	.23161	.72762	.19867	1.9280	.64372	.37694	.35080	.68663
#1	.09280	.23499	.29694	1.9564	29.779	.51781	72.039	.61218
#2	.09315	.23843	.29739	1.9005	29.610	.51588	71.921	.60495
#3	.09276	.23672	.29622	1.8867	29.399	.51393	71.556	.60497
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49958**	**89.482**	**.23320**	**4.8335**	**.23655**	**.57885**	**.18210**	**6.3388**
Stddev	.00153	.784	.00050	.0076	.00169	.00511	.00473	.0094
%RSD	.30585	.87569	.21296	.15803	.71424	.88252	2.5997	.14817
#1	.49951	90.366	.23335	4.8331	.23580	.57825	.18575	6.3355
#2	.50114	89.209	.23361	4.8414	.23849	.57407	.18379	6.3494
#3	.49809	88.872	.23265	4.8261	.23537	.58424	.17675	6.3314
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098636SD Acquired: 6/25/2016 19:13:21 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-08

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47318**	**2.1504**	**.45196**	**.23380**	**.49243**	**.47726**	**.00396**
Stddev	.00235	.0203	.01296	.00229	.00254	.00239	.00140
%RSD	.49768	.94209	2.8683	.97909	.51630	.50171	35.414
#1	.47220	2.1736	.45301	.23117	.49302	.48002	.00408
#2	.47587	2.1417	.46436	.23534	.49462	.47583	.00530
#3	.47148	2.1359	.43850	.23489	.48964	.47592	.00250
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9466.5**	**57687.**	**3000.7**
Stddev	28.5	42.	21.4
%RSD	.30105	.07212	.71153
#1	9499.3	57706.	2976.5
#2	9448.1	57639.	3017.0
#3	9452.0	57716.	3008.4

Approved: June 27, 2016

Sample Name: L1606098638SD Acquired: 6/25/2016 19:16:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-09

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19635**	**5.0111**	**.19662**	**1.3658**	**.50590**	**.02455**	**238.74**	**.02470**
Stddev	.00189	.0495	.00202	.0129	.00507	.00013	2.44	.00046
%RSD	.96164	.98714	1.0288	.94297	1.0027	.51871	1.0237	1.8632
#1	.19421	4.9540	.19875	1.3513	.50004	.02440	236.05	.02465
#2	.19777	5.0396	.19638	1.3761	.50894	.02464	240.82	.02518
#3	.19706	5.0398	.19472	1.3699	.50872	.02461	239.35	.02426
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09456**	**.24315**	**.24888**	**1.9280**	**29.506**	**.51676**	**69.466**	**.60260**
Stddev	.00054	.00286	.00145	.0328	.362	.00473	.757	.00806
%RSD	.56749	1.1744	.58163	1.6990	1.2271	.91552	1.0898	1.3377
#1	.09518	.24020	.24797	1.8914	29.119	.51186	68.628	.59529
#2	.09432	.24590	.25055	1.9547	29.837	.52131	70.101	.61124
#3	.09419	.24336	.24812	1.9378	29.562	.51712	69.668	.60126
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50828**	**86.839**	**.23731**	**4.9079**	**.23796**	**.58722**	**.18727**	**6.3571**
Stddev	.00153	.812	.00040	.0105	.00122	.00763	.00717	.0147
%RSD	.30188	.93521	.16702	.21476	.51147	1.2987	3.8265	.23138
#1	.50998	85.938	.23758	4.9185	.23889	.59365	.18466	6.3682
#2	.50784	87.516	.23685	4.9076	.23841	.58922	.19537	6.3627
#3	.50701	87.061	.23748	4.8975	.23659	.57879	.18177	6.3405
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098638SD Acquired: 6/25/2016 19:16:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-09

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48230**	**2.1413**	**.45206**	**.24148**	**.50819**	**.47197**	**.00465**
Stddev	.00016	.0221	.00841	.00203	.00480	.00162	.00175
%RSD	.03337	1.0323	1.8614	.84256	.94539	.34380	37.692
#1	.48228	2.1164	.44367	.23931	.50264	.47335	.00633
#2	.48215	2.1586	.45201	.24177	.51108	.47237	.00283
#3	.48247	2.1488	.46050	.24334	.51084	.47018	.00481
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9367.1**	**56617.**	**3005.4**
Stddev	24.2	279.	8.9
%RSD	.25886	.49353	.29743
#1	9345.0	56900.	3015.6
#2	9363.2	56609.	2998.7
#3	9393.0	56341.	3002.0

Approved: June 27, 2016

Sample Name: L1606098640 Acquired: 6/25/2016 19:20:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00070	.00014	.00047	.01204	.00394	.00013	2.2418	.00009
Stddev	.00052	.00246	.00354	.00206	.00041	.00015	.0448	.00026
%RSD	75.327	1722.8	753.06	17.114	10.437	111.96	1.9977	298.05
#1	-.00120	.00298	.00434	.01328	.00360	.00001	2.2906	-.00017
#2	-.00073	-.00107	-.00032	.00966	.00384	.00009	2.2025	.00008
#3	-.00016	-.00148	-.00261	.01318	.00440	.00029	2.2324	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00011	-.00023	.00162	.03197	.19488	-.00199	.70657	-.00157
Stddev	.00020	.00043	.00063	.03003	.05117	.00159	.10487	.00034
%RSD	172.28	188.43	38.822	93.912	26.255	79.698	14.843	21.723
#1	-.00014	.00024	.00129	.06510	.20250	-.00219	.77648	-.00118
#2	.00010	-.00061	.00122	.00654	.14033	-.00346	.58598	-.00170
#3	-.00030	-.00032	.00234	.02429	.24180	-.00031	.75725	-.00182
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00061	1.3245	-.00000	-.00132	.00155	-.00034	.00103	1.8661
Stddev	.00043	.0047	.00128	.00539	.00283	.00334	.00522	.0030
%RSD	69.622	.35334	42382.	408.47	183.18	979.63	508.03	.16155
#1	.00110	1.3264	-.00135	.00378	-.00057	-.00417	-.00062	1.8696
#2	.00029	1.3279	.00121	-.00695	.00477	.00193	.00687	1.8648
#3	.00045	1.3192	.00013	-.00078	.00045	.00122	-.00317	1.8641
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098640 Acquired: 6/25/2016 19:20:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00060	.02890	-.00018	-.00245	.00125	.00316	.00092
Stddev	.00104	.00022	.00313	.00122	.00036	.00008	.00187
%RSD	172.87	.75581	1755.3	49.848	28.928	2.5446	202.44
#1	-.00053	.02865	.00335	-.00105	.00165	.00322	-.00102
#2	.00152	.02900	-.00128	-.00328	.00095	.00320	.00271
#3	.00082	.02905	-.00260	-.00302	.00115	.00307	.00108
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9902.6	60698.	3047.4
Stddev	11.3	88.	21.1
%RSD	.11376	.14441	.69149
#1	9915.6	60659.	3066.8
#2	9895.9	60798.	3025.0
#3	9896.3	60636.	3050.3

Approved: June 27, 2016

Sample Name: L1606098642 Acquired: 6/25/2016 19:24:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00084	-.00125	.00012	.00714	.00291	.00014	2.2957	-.00008
Stddev	.00151	.00416	.00315	.00149	.00136	.00010	.0287	.00011
%RSD	179.62	331.72	2535.0	20.853	46.790	76.580	1.2503	140.66
#1	.00086	.00295	-.00350	.00543	.00443	.00023	2.2645	-.00020
#2	-.00134	-.00134	.00221	.00806	.00180	.00015	2.3210	.00001
#3	-.00204	-.00537	.00167	.00794	.00250	.00002	2.3016	-.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00027	.00054	.00126	.01879	.13459	.00101	.70251	-.00031
Stddev	.00021	.00073	.00033	.01779	.06299	.00139	.04015	.00188
%RSD	76.216	135.58	26.342	94.669	46.805	137.70	5.7150	613.21
#1	.00043	.00109	.00162	.03687	.13442	.00244	.65916	.00151
#2	.00035	-.00029	.00096	.01820	.19766	.00092	.73842	-.00224
#3	.00004	.00081	.00121	.00130	.07167	-.00034	.70994	-.00020
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00012	1.3914	-.00012	.00599	-.00050	-.00197	.00192	1.9147
Stddev	.00031	.0274	.00128	.00446	.00248	.00478	.00405	.0013
%RSD	265.01	1.9704	1039.6	74.477	491.66	243.04	211.10	.06740
#1	-.00017	1.4067	.00078	.00869	.00158	.00334	.00542	1.9141
#2	.00008	1.4078	.00044	.00843	.00015	-.00331	-.00251	1.9137
#3	.00044	1.3598	-.00159	.00084	-.00325	-.00593	.00284	1.9161
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098642 Acquired: 6/25/2016 19:24:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**.02978**	**.00659**	**-.00224**	**.00021**	**.00288**	**.00046**
Stddev	.00041	.00031	.00898	.00218	.00024	.00011	.00203
%RSD	319.77	1.0309	136.22	97.103	115.48	3.6974	442.65
#1	.00034	.02969	.01094	-.00308	.00030	.00286	-.00018
#2	-.00043	.03012	.01258	.00023	.00040	.00300	.00273
#3	-.00030	.02953	-.00373	-.00388	-.00006	.00279	-.00117
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9910.0**	**60405.**	**3027.1**
Stddev	8.6	109.	25.1
%RSD	.08719	.18086	.83024
#1	9909.0	60400.	3001.0
#2	9919.1	60298.	3029.3
#3	9901.9	60516.	3051.1

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 19:28:27 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.38218**	**9.6308**	**.37557**	**.46431**	**.95962**	**.04828**	**9.5695**
Stddev	.00218	.0178	.00127	.00116	.00217	.00012	.0717
%RSD	.57165	.18471	.33835	.24921	.22635	.23895	.74942
#1	.37992	9.6388	.37606	.46299	.96032	.04841	9.5557
#2	.38232	9.6104	.37652	.46516	.96137	.04820	9.6471
#3	.38429	9.6432	.37413	.46479	.95719	.04823	9.5057
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.04801**	**.18801**	**.47395**	**.48473**	**3.7689**	**46.355**	**.97311**
Stddev	.00014	.00020	.00045	.00153	.0264	.220	.00378
%RSD	.29427	.10445	.09483	.31643	.70136	.47512	.38849
#1	.04785	.18823	.47438	.48634	3.7952	46.500	.97746
#2	.04812	.18785	.47399	.48328	3.7693	46.464	.97068
#3	.04806	.18795	.47348	.48458	3.7423	46.102	.97118
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.0717**	**.47590**	**.96524**	**47.107**	**.47852**	**9.3733**	**.47882**
Stddev	.0608	.00396	.00399	.173	.00072	.0111	.00296
%RSD	.67059	.83199	.41378	.36832	.15104	.11884	.61810
#1	9.0038	.47335	.96959	47.188	.47773	9.3765	.47551
#2	9.1213	.48046	.96440	47.226	.47916	9.3609	.48123
#3	9.0900	.47388	.96173	46.908	.47866	9.3825	.47971
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 19:28:27 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1223**	**.37795**	**4.8400**	**.94463**	**.95969**	**.96684**	**.48966**
Stddev	.0016	.00541	.0100	.00151	.00470	.01176	.00307
%RSD	.14404	1.4313	.20631	.15941	.49016	1.2167	.62598
#1	1.1238	.37410	4.8492	.94636	.96229	.96908	.48613
#2	1.1206	.37562	4.8414	.94360	.96253	.97732	.49169
#3	1.1224	.38414	4.8294	.94394	.95426	.95412	.49116
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.97958**	**.95235**	F **2.8246**
Stddev	.00154	.00042	.0035
%RSD	.15724	.04377	.12384
#1	.97882	.95263	2.8278
#2	.97857	.95254	2.8252
#3	.98136	.95187	2.8208
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10132.**	**61264.**	**3100.4**
Stddev	.	134.	6.9
%RSD	.00202	.21941	.22117
#1	10132.	61202.	3092.6
#2	10132.	61418.	3103.3
#3	10132.	61171.	3105.3

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 19:32:05 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00027**	**-.00034**	**.00165**	**-.00065**	**.00143**	**.00014**	**-.04509**	**-.00015**
Stddev	.00053	.00835	.00116	.00135	.00075	.00012	.04632	.00018
%RSD	198.69	2437.1	70.395	208.11	52.425	84.191	102.74	120.05
#1	.00081	.00499	.00229	-.00108	.00058	.00001	-.09453	-.00022
#2	.00023	-.00997	.00235	-.00174	.00173	.00020	-.03804	.00005
#3	-.00025	.00395	.00031	.00087	.00199	.00023	-.00270	-.00028
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00008**	**-.00017**	**.00065**	**.02423**	**-.06433**	**-.00295**	**-.01903**	**-.00066**
Stddev	.00013	.00077	.00059	.01555	.07126	.00237	.08401	.00260
%RSD	162.10	456.13	90.441	64.178	110.77	80.373	441.41	395.33
#1	-.00022	-.00087	-.00002	.03897	.01654	-.00026	-.10302	.00036
#2	-.00006	.00066	.00091	.02575	-.09166	-.00387	.06499	-.00361
#3	.00004	-.00030	.00106	.00798	-.11789	-.00472	-.01907	.00127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00313**	**.02145**	**.00120**	**.00197**	**.00164**	**.00212**	**.00101**	**.04112**
Stddev	.00013	.00743	.00050	.00366	.00312	.00166	.00069	.00272
%RSD	4.1237	34.660	41.579	185.66	190.04	78.472	68.223	6.6255
#1	.00318	.03002	.00170	.00620	.00310	.00128	.00172	.03946
#2	.00323	.01754	.00070	-.00004	.00377	.00104	.00035	.04426
#3	.00298	.01678	.00121	-.00024	-.00194	.00404	.00097	.03964
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 19:32:05 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00072**	**.00051**	**-.00006**	**-.00149**	**-.00026**	**.00006**	**.00316**
Stddev	.00004	.00041	.00466	.00123	.00016	.00032	.00144
%RSD	4.9365	80.523	8195.1	82.501	61.089	521.42	45.405
#1	.00073	.00010	.00498	-.00182	-.00044	-.00016	.00318
#2	.00068	.00091	-.00094	-.00252	-.00020	.00042	.00172
#3	.00075	.00050	-.00421	-.00013	-.00014	-.00008	.00459
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9625.9**	**58446.**	**2947.7**
Stddev	17.4	102.	8.2
%RSD	.18036	.17465	.27941
#1	9605.9	58491.	2954.6
#2	9635.9	58518.	2950.0
#3	9636.0	58329.	2938.6

Approved: June 27, 2016

Sample Name: PBW LH Acquired: 6/25/2016 19:36:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00022	-.00480	.00297	-.00088	.00071	.00016	-.03609	-.00030
Stddev	.00096	.00538	.00061	.00047	.00081	.00006	.00576	.00022
%RSD	436.54	112.05	20.402	53.250	112.94	41.040	15.975	74.291
#1	-.00063	-.00271	.00258	-.00071	.00057	.00010	-.04014	-.00038
#2	.00002	-.00078	.00267	-.00141	-.00001	.00022	-.03863	-.00005
#3	.00126	-.01092	.00367	-.00052	.00158	.00014	-.02949	-.00047
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00009	.00027	.00021	.03551	.02057	.00058	-.03422	.00199
Stddev	.00019	.00046	.00040	.00062	.11033	.00127	.05831	.00213
%RSD	211.66	173.58	192.44	1.7348	536.36	218.26	170.39	107.33
#1	.00009	.00043	-.00014	.03559	.14793	.00205	.01703	.00051
#2	-.00029	.00063	.00065	.03485	-.04548	-.00019	-.09766	.00101
#3	-.00007	-.00025	.00012	.03607	-.04075	-.00011	-.02203	.00443
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00027	.01920	-.00013	.00015	.00308	-.00169	.00218	.00530
Stddev	.00020	.01381	.00043	.00436	.00410	.00157	.00344	.00135
%RSD	72.649	71.919	323.98	2952.6	133.25	92.573	157.59	25.417
#1	-.00050	.00540	-.00047	-.00185	.00268	-.00008	.00514	.00605
#2	-.00014	.01919	.00035	-.00286	.00737	-.00321	.00301	.00374
#3	-.00017	.03301	-.00027	.00515	-.00081	-.00179	-.00160	.00610
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: PBW LH Acquired: 6/25/2016 19:36:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00038**	**.00001**	**.00007**	**-.00205**	**.00056**	**-.00012**	**.00138**
Stddev	.00105	.00006	.00526	.00274	.00041	.00008	.00122
%RSD	278.35	503.89	7708.4	133.89	73.183	68.172	88.458
#1	.00057	-.00005	.00045	-.00401	.00011	-.00018	.00093
#2	-.00020	.00002	-.00537	-.00321	.00065	-.00003	.00045
#3	-.00151	.00007	.00513	.00108	.00091	-.00016	.00276
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10355.**	**62690.**	**3143.1**
Stddev	14.	21.	40.1
%RSD	.13810	.03356	1.2743
#1	10371.	62711.	3176.8
#2	10343.	62669.	3098.8
#3	10351.	62689.	3153.6

Approved: June 27, 2016

Sample Name: LCSW LH Acquired: 6/25/2016 19:39:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19366**	**4.9137**	**.19220**	**.93203**	**.49068**	**.02467**	**4.9632**	**.02443**
Stddev	.00070	.0138	.00341	.00396	.00423	.00007	.0538	.00016
%RSD	.36335	.28142	1.7740	.42472	.86189	.28277	1.0835	.64799
#1	.19288	4.8983	.19002	.92766	.48616	.02461	4.9413	.02459
#2	.19385	4.9180	.19044	.93305	.49453	.02465	5.0244	.02441
#3	.19425	4.9249	.19613	.93537	.49136	.02474	4.9238	.02428
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09715**	**.24544**	**.24999**	**1.9443**	**23.472**	**.50293**	**4.6813**	**.24132**
Stddev	.00066	.00263	.00041	.0106	.182	.00291	.0938	.00318
%RSD	.68088	1.0726	.16565	.54556	.77586	.57762	2.0038	1.3188
#1	.09680	.24301	.25039	1.9373	23.301	.50365	4.6704	.23802
#2	.09792	.24824	.24956	1.9391	23.664	.50540	4.7800	.24157
#3	.09674	.24506	.25002	1.9565	23.450	.49973	4.5934	.24437
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50304**	**24.148**	**.24571**	**4.6828**	**.24554**	**.57884**	**.18202**	**2.4064**
Stddev	.00049	.122	.00061	.0125	.00123	.00461	.00367	.0046
%RSD	.09743	.50324	.24625	.26789	.50266	.79649	2.0171	.19316
#1	.50339	24.028	.24640	4.6758	.24587	.57352	.17868	2.4078
#2	.50325	24.271	.24545	4.6973	.24658	.58167	.18595	2.4102
#3	.50248	24.145	.24528	4.6753	.24418	.58132	.18143	2.4012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LCSW LH Acquired: 6/25/2016 19:39:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48493**	**.49684**	**.48674**	**.24907**	**.50464**	**.48284**	**.01676**
Stddev	.00108	.00397	.00310	.00041	.00236	.00325	.00211
%RSD	.22287	.79928	.63744	.16637	.46686	.67256	12.564
#1	.48462	.49378	.48665	.24947	.50200	.48055	.01435
#2	.48613	.50133	.48369	.24909	.50538	.48656	.01775
#3	.48404	.49542	.48989	.24864	.50653	.48141	.01820
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9761.3**	**59171.**	**3026.9**
Stddev	5.3	58.	23.7
%RSD	.05400	.09751	.78195
#1	9766.4	59236.	3050.7
#2	9755.9	59148.	3003.3
#3	9761.7	59127.	3026.7

Approved: June 27, 2016

Sample Name: L1606090723 Acquired: 6/25/2016 19:43:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00091	.01800	.14333	.20159	.04322	.00011	82.135	.00062
Stddev	.00074	.00440	.00262	.00108	.00176	.00009	.153	.00002
%RSD	81.113	24.457	1.8284	.53531	4.0750	85.790	.18660	3.6806
#1	.00039	.02274	.14560	.20094	.04396	.00020	82.043	.00061
#2	.00058	.01723	.14046	.20283	.04121	.00011	82.051	.00064
#3	.00176	.01403	.14393	.20099	.04450	.00001	82.312	.00060
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00085	.00186	.00181	3.0759	4.3146	.03826	51.383	.03252
Stddev	.00009	.00076	.00070	.0448	.0921	.00398	.083	.00353
%RSD	10.346	41.113	38.683	1.4576	2.1337	10.401	.16137	10.870
#1	.00081	.00098	.00242	3.0253	4.2316	.04144	51.475	.03488
#2	.00094	.00223	.00197	3.0916	4.4136	.03380	51.362	.02845
#3	.00078	.00237	.00105	3.1108	4.2984	.03956	51.313	.03422
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00258	133.73	-.00015	.01191	.00179	-.00228	.00665	8.0891
Stddev	.00008	.12	.00113	.00694	.00129	.00283	.00532	.0280
%RSD	3.2153	.09189	747.26	58.268	71.933	124.12	79.994	.34670
#1	.00264	133.81	-.00126	.01916	.00318	-.00553	.01195	8.1130
#2	.00249	133.59	.00099	.00533	.00159	-.00033	.00671	8.0582
#3	.00262	133.80	-.00018	.01123	.00062	-.00099	.00131	8.0960
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606090723 Acquired: 6/25/2016 19:43:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00001**	**2.6416**	**-.00732**	**-.00182**	**-.00046**	**.00542**	**.00128**
Stddev	.00038	.0023	.00771	.00082	.00062	.00021	.00019
%RSD	7616.2	.08543	105.44	45.025	136.48	3.8805	14.578
#1	.00020	2.6400	-.01112	-.00260	-.00073	.00544	.00120
#2	.00025	2.6406	.00156	-.00189	.00026	.00563	.00115
#3	-.00044	2.6441	-.01239	-.00097	-.00089	.00521	.00149
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9566.3**	**57706.**	**3016.3**
Stddev	21.2	40.	8.6
%RSD	.22170	.06981	.28561
#1	9583.7	57737.	3006.9
#2	9572.5	57660.	3023.9
#3	9542.7	57721.	3018.2

Approved: June 27, 2016

Sample Name: L1606098601 Acquired: 6/25/2016 19:47:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00061	.63226	.00101	.09520	.04126	.00012	31.325	.00014
Stddev	.00183	.00734	.00119	.00196	.00050	.00007	.135	.00021
%RSD	298.65	1.1612	118.08	2.0630	1.2197	57.984	.43030	153.25
#1	.00185	.62395	-.00028	.09744	.04176	.00005	31.451	-.00001
#2	-.00149	.63786	.00207	.09436	.04076	.00018	31.183	.00005
#3	.00148	.63498	.00124	.09379	.04125	.00013	31.342	.00038
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00117	.00251	.00378	1.4621	1.2075	.01677	5.9028	.04892
Stddev	.00024	.00020	.00199	.0362	.0294	.00358	.0605	.00066
%RSD	20.175	7.8011	52.767	2.4785	2.4351	21.330	1.0243	1.3538
#1	.00129	.00246	.00338	1.4981	1.1821	.01735	5.9016	.04861
#2	.00090	.00272	.00594	1.4256	1.2008	.02003	5.8430	.04848
#3	.00133	.00234	.00201	1.4624	1.2397	.01294	5.9639	.04969
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00119	122.45	.00122	.03952	.00058	-.00189	.00677	4.0503
Stddev	.00024	.35	.00124	.00581	.00157	.00112	.00463	.0371
%RSD	19.925	.28250	101.52	14.694	270.52	59.209	68.294	.91471
#1	.00135	122.68	.00181	.04077	.00200	-.00256	.00933	4.0789
#2	.00131	122.05	-.00020	.04460	-.00110	-.00252	.00956	4.0634
#3	.00092	122.62	.00206	.03319	.00084	-.00060	.00143	4.0084
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098601 Acquired: 6/25/2016 19:47:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00016**	**.54428**	**.01566**	**-.00252**	**.00147**	**.03503**	**.00138**
Stddev	.00052	.00100	.00262	.00158	.00115	.00040	.00180
%RSD	330.58	.18426	16.757	62.766	78.497	1.1335	130.15
#1	-.00032	.54517	.01814	-.00107	.00112	.03533	.00331
#2	-.00057	.54319	.01592	-.00228	.00275	.03517	-.00024
#3	.00042	.54448	.01291	-.00422	.00053	.03458	.00107
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9674.1**	**58435.**	**3018.4**
Stddev	45.2	129.	18.2
%RSD	.46720	.22152	.60258
#1	9673.0	58340.	2997.4
#2	9719.9	58382.	3028.7
#3	9629.5	58582.	3029.1

Approved: June 27, 2016

Sample Name: L1606098603 Acquired: 6/25/2016 19:51:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00113**	**.08091**	**.00257**	**.02612**	**.02343**	**.00026**	**258.72**
Stddev	.00364	.00403	.00338	.00065	.00023	.00010	.22
%RSD	322.73	4.9793	131.60	2.5002	.99712	37.911	.08495
#1	-.00201	.07953	.00151	.02617	.02317	.00019	258.95
#2	.00512	.08545	-.00015	.02675	.02363	.00037	258.52
#3	.00028	.07776	.00635	.02545	.02348	.00021	258.68
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00056**	**.00486**	**.00101**	**.01760**	**1.6277**	**1.7887**	**.01926**
Stddev	.00018	.00049	.00101	.00048	.0214	.0748	.00269
%RSD	32.935	10.099	100.65	2.7241	1.3157	4.1799	13.963
#1	.00053	.00446	.00070	.01814	1.6519	1.7719	.02232
#2	.00040	.00472	.00018	.01725	1.6198	1.8705	.01729
#3	.00076	.00541	.00214	.01740	1.6113	1.7238	.01816
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**67.571**	**3.7591**	**.00091**	**16.176**	**.00450**	**.03792**	**.00200**
Stddev	.265	.0091	.00013	.042	.00050	.00459	.00102
%RSD	.39176	.24313	14.640	.25963	11.081	12.099	51.049
#1	67.875	3.7689	.00083	16.148	.00428	.03292	.00317
#2	67.391	3.7508	.00083	16.224	.00415	.03891	.00158
#3	67.448	3.7575	.00106	16.157	.00507	.04193	.00126
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: L1606098603 Acquired: 6/25/2016 19:51:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00201**	**-.00717**	**3.9929**	**.00071**	**1.4034**	F **-.03512**	**-.00302**
Stddev	.00033	.00269	.0187	.00006	.0013	.00376	.00159
%RSD	16.650	37.521	.46951	8.6416	.08919	10.706	52.853
#1	-.00201	-.00583	3.9724	.00069	1.4035	-.03701	-.00329
#2	-.00234	-.01026	3.9969	.00078	1.4046	-.03756	-.00130
#3	-.00167	-.00541	4.0093	.00066	1.4021	-.03079	-.00446
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.00011**	**.03959**	**.00096**
Stddev	.00028	.00012	.00361
%RSD	248.38	.29182	376.14
#1	.00034	.03947	.00158
#2	-.00020	.03969	.00421
#3	.00020	.03962	-.00292
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9451.5**	**57637.**	**3009.6**
Stddev	22.8	186.	22.0
%RSD	.24088	.32311	.72999
#1	9477.8	57852.	2986.4
#2	9439.3	57520.	3012.4
#3	9437.5	57540.	3030.0

Approved: June 27, 2016

Sample Name: L1606098605 Acquired: 6/25/2016 19:55:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00137	.11904	-.00057	.01716	.07339	.00069	79.533	.00007
Stddev	.00111	.00237	.00183	.00141	.00107	.00010	.468	.00014
%RSD	81.063	1.9908	322.69	8.2448	1.4582	14.460	.58849	215.58
#1	-.00160	.12158	.00130	.01581	.07215	.00080	78.994	-.00002
#2	-.00016	.11865	-.00237	.01703	.07401	.00066	79.770	-.00001
#3	-.00235	.11689	-.00063	.01863	.07401	.00061	79.836	.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00018	.00197	.03453	.22031	1.0996	.00600	23.636	.02412
Stddev	.00010	.00062	.00164	.01719	.0321	.00378	.124	.00114
%RSD	58.539	31.534	4.7611	7.8016	2.9198	62.984	.52590	4.7055
#1	.00028	.00148	.03642	.22738	1.1284	.00174	23.677	.02484
#2	.00008	.00176	.03343	.23283	1.0650	.00731	23.734	.02281
#3	.00016	.00266	.03374	.20072	1.1054	.00895	23.496	.02471
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00088	16.543	.00139	.02145	.00230	-.00291	.00287	2.5906
Stddev	.00010	.077	.00029	.00514	.00167	.00113	.00378	.0043
%RSD	11.067	.46265	20.622	23.950	72.689	38.933	132.03	.16499
#1	.00094	16.521	.00147	.02583	.00422	-.00367	.00465	2.5950
#2	.00093	16.628	.00163	.02271	.00152	-.00161	-.00148	2.5865
#3	.00076	16.480	.00107	.01580	.00116	-.00345	.00543	2.5902
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098605 Acquired: 6/25/2016 19:55:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00022**	**.41216**	**-.00457**	**-.00401**	**.00095**	**.08292**	**.00137**
Stddev	.00045	.00253	.00113	.00092	.00095	.00025	.00112
%RSD	198.91	.61368	24.692	22.922	99.932	.30169	82.337
#1	.00020	.40988	-.00328	-.00486	-.00009	.08312	.00008
#2	.00068	.41488	-.00504	-.00303	.00116	.08301	.00218
#3	-.00021	.41173	-.00539	-.00414	.00177	.08264	.00183
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9691.4**	**59019.**	**3013.1**
Stddev	15.7	113.	12.9
%RSD	.16185	.19156	.42939
#1	9673.9	58889.	3028.0
#2	9704.2	59073.	3004.1
#3	9696.2	59095.	3007.4

Approved: June 27, 2016

Sample Name: L1606098607 Acquired: 6/25/2016 19:59:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00078	.03854	.00139	.01595	.03639	.00021	86.351	.00013
Stddev	.00196	.00438	.00171	.00299	.00069	.00005	.183	.00018
%RSD	252.38	11.356	122.54	18.743	1.8946	25.559	.21179	130.52
#1	-.00064	.04027	.00252	.01429	.03564	.00026	86.561	.00033
#2	-.00004	.03356	-.00057	.01941	.03654	.00022	86.226	.00007
#3	.00302	.04179	.00223	.01417	.03700	.00015	86.267	-.00000
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00086	.00096	.00901	.09595	1.3588	.01896	23.947	.13765
Stddev	.00012	.00155	.00054	.00336	.0905	.00285	.152	.00156
%RSD	14.271	161.05	6.0176	3.5057	6.6619	15.032	.63449	1.1314
#1	.00074	.00257	.00840	.09606	1.2544	.01772	23.782	.13585
#2	.00084	-.00052	.00942	.09926	1.4148	.02223	24.081	.13859
#3	.00099	.00083	.00922	.09254	1.4072	.01695	23.978	.13851
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	12.868	.00006	.00776	.00294	-.00141	.00160	3.6487
Stddev	.00036	.033	.00069	.00647	.00245	.00292	.00413	.0140
%RSD	74.294	.25857	1239.7	83.320	83.569	207.44	258.10	.38292
#1	.00008	12.905	-.00031	.01172	.00013	.00158	.00626	3.6621
#2	.00078	12.859	.00085	.01126	.00466	-.00154	-.00162	3.6496
#3	.00059	12.840	-.00038	.00030	.00403	-.00426	.00017	3.6342
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098607 Acquired: 6/25/2016 19:59:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**.85925**	**-.01297**	**-.00261**	**.00027**	**.01872**	**-.00023**
Stddev	.00058	.00368	.00482	.00118	.00049	.00005	.00100
%RSD	1135.5	.42821	37.150	45.319	182.28	.27976	440.95
#1	.00028	.86236	-.00802	-.00398	-.00019	.01878	.00007
#2	-.00072	.86019	-.01323	-.00183	.00021	.01871	.00059
#3	.00028	.85519	-.01765	-.00203	.00079	.01868	-.00134
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9667.2**	**58651.**	**3007.8**
Stddev	38.0	90.	17.4
%RSD	.39269	.15372	.57760
#1	9636.2	58610.	2989.6
#2	9655.8	58589.	3009.4
#3	9709.5	58755.	3024.3

Approved: June 27, 2016

Sample Name: L1606098607PS Acquired: 6/25/2016 20:03:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG574215-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19549**	**4.9683**	**.19699**	**.95431**	**.52192**	**.02489**	**83.384**	**.02448**
Stddev	.00145	.0012	.00037	.00069	.00335	.00004	.351	.00015
%RSD	.74419	.02324	.18968	.07193	.64144	.17436	.42141	.59448
#1	.19431	4.9671	.19669	.95367	.51889	.02485	83.062	.02455
#2	.19712	4.9693	.19740	.95504	.52551	.02494	83.759	.02458
#3	.19505	4.9687	.19686	.95422	.52137	.02490	83.330	.02431
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09612**	**.24445**	**.24901**	**2.0095**	**25.021**	**.52274**	**26.551**	**.36466**
Stddev	.00060	.00204	.00132	.0079	.209	.00262	.211	.00375
%RSD	.62715	.83504	.52860	.39401	.83514	.50106	.79467	1.0287
#1	.09620	.24322	.24870	2.0100	24.845	.52124	26.458	.36188
#2	.09668	.24681	.25046	2.0172	25.252	.52577	26.792	.36317
#3	.09548	.24332	.24788	2.0014	24.965	.52123	26.403	.36893
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50539**	**35.868**	**.24056**	**4.8300**	**.24330**	**.58213**	**.19141**	**5.6770**
Stddev	.00256	.121	.00174	.0165	.00225	.00117	.00102	.0113
%RSD	.50685	.33731	.72421	.34266	.92572	.20086	.53487	.19978
#1	.50625	35.730	.24173	4.8368	.24292	.58281	.19254	5.6837
#2	.50742	35.956	.24138	4.8421	.24572	.58078	.19056	5.6833
#3	.50251	35.919	.23856	4.8112	.24127	.58280	.19112	5.6639
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098607PS Acquired: 6/25/2016 20:03:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG574215-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48483**	**1.2623**	**.48068**	**.24615**	**.50488**	**.48944**	**.00427**
Stddev	.00147	.0060	.00508	.00275	.00238	.00115	.00064
%RSD	.30392	.47569	1.0573	1.1154	.47121	.23489	15.068
#1	.48451	1.2555	.47807	.24920	.50216	.48999	.00389
#2	.48644	1.2668	.48654	.24388	.50587	.49021	.00501
#3	.48354	1.2646	.47744	.24537	.50660	.48812	.00390
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9610.5**	**58019.**	**3004.1**
Stddev	20.4	180.	8.2
%RSD	.21278	.31055	.27313
#1	9607.7	58202.	3005.9
#2	9591.6	58012.	2995.2
#3	9632.2	57842.	3011.3

Approved: June 27, 2016

Sample Name: L1606098607SDL Acquired: 6/25/2016 20:06:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG574215-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00055	.00334	.00170	.00445	.00847	.00009	17.139	.00008
Stddev	.00049	.00730	.00411	.00207	.00008	.00007	.249	.00023
%RSD	88.649	218.69	241.70	46.624	.94243	75.660	1.4519	290.72
#1	.00089	.00593	.00612	.00219	.00843	.00006	17.426	.00034
#2	-.00001	.00899	-.00200	.00488	.00843	.00004	17.001	-.00003
#3	.00078	-.00490	.00098	.00627	.00857	.00017	16.989	-.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00056	.00037	.00191	.01894	.32539	.00400	4.9234	.02615
Stddev	.00005	.00023	.00083	.00475	.10069	.00194	.1236	.00303
%RSD	9.1063	60.795	43.755	25.087	30.943	48.604	2.5107	11.568
#1	.00062	.00059	.00284	.02402	.21108	.00176	4.9976	.02622
#2	.00053	.00038	.00165	.01821	.40094	.00496	4.9918	.02914
#3	.00053	.00014	.00123	.01460	.36416	.00527	4.7807	.02309
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00031	2.5977	.00028	.00852	.00297	.00160	.00156	.72149
Stddev	.00021	.0380	.00123	.00414	.00293	.00068	.00366	.00246
%RSD	66.333	1.4636	438.92	48.637	98.645	42.261	235.25	.34147
#1	.00008	2.6376	-.00081	.01268	-.00028	.00113	-.00104	.72244
#2	.00048	2.5938	.00004	.00439	.00377	.00130	.00574	.71869
#3	.00037	2.5619	.00161	.00849	.00542	.00238	-.00003	.72333
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098607SDL Acquired: 6/25/2016 20:06:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG574215-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**.17255**	**.00033**	**-.00095**	**.00100**	**.00419**	**.00060**
Stddev	.00072	.00063	.00618	.00123	.00055	.00013	.00060
%RSD	1344.6	.36773	1857.8	129.31	55.083	2.9985	99.890
#1	-.00060	.17322	-.00395	-.00224	.00116	.00417	-.00009
#2	.00076	.17246	-.00247	-.00083	.00039	.00432	.00094
#3	-.00032	.17196	.00742	.00021	.00145	.00407	.00094
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10100.**	**61152.**	**3087.7**
Stddev	17.	221.	3.3
%RSD	.16887	.36180	.10673
#1	10112.	61407.	3088.4
#2	10080.	61005.	3090.6
#3	10107.	61046.	3084.1

Approved: June 27, 2016

Sample Name: L1606098607SDL Acquired: 6/25/2016 20:10:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG574215-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00016**	**-.00220**	**-.00032**	**-.00063**	**.00197**	**.00015**	**3.3784**	**-.00009**
Stddev	.00281	.00599	.00141	.00122	.00070	.00004	.0403	.00029
%RSD	1738.4	271.64	438.98	193.96	35.714	23.124	1.1918	310.09
#1	.00289	-.00340	.00101	-.00187	.00220	.00018	3.3321	-.00013
#2	-.00273	.00429	-.00181	-.00058	.00253	.00017	3.3977	-.00037
#3	.00033	-.00751	-.00017	.00056	.00118	.00011	3.4054	.00021
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00001**	**-.00057**	**.00039**	**.03496**	**.02799**	**.00189**	**.96584**	**.00502**
Stddev	.00041	.00109	.00122	.02559	.02173	.00316	.06365	.00248
%RSD	5673.0	191.57	314.78	73.189	77.636	167.39	6.5899	49.517
#1	.00017	-.00161	.00178	.00957	.03029	.00425	1.0133	.00340
#2	.00028	-.00067	-.00011	.06074	.04847	.00313	.99069	.00377
#3	-.00047	.00057	-.00050	.03458	.00520	-.00171	.89352	.00788
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00037**	**.48659**	**-.00001**	**.00389**	**.00194**	**-.00589**	**.00123**	**.14706**
Stddev	.00019	.02107	.00027	.00257	.00032	.00198	.00158	.00411
%RSD	50.403	4.3293	3257.2	66.021	16.675	33.653	128.28	2.7928
#1	-.00035	.50181	-.00003	.00472	.00177	-.00726	-.00048	.14268
#2	-.00056	.46255	.00027	.00101	.00173	-.00680	.00263	.15082
#3	-.00019	.49542	-.00027	.00595	.00231	-.00362	.00154	.14770
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098607SDL Acquired: 6/25/2016 20:10:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG574215-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00022**	**.03404**	**-.00096**	**.00109**	**.00053**	**.00095**	**-.00050**
Stddev	.00021	.00026	.00497	.00270	.00028	.00007	.00147
%RSD	93.819	.76060	519.10	248.53	53.226	6.9237	292.99
#1	-.00042	.03414	.00415	-.00145	.00059	.00088	.00090
#2	-.00001	.03374	-.00578	.00079	.00078	.00097	-.00204
#3	-.00022	.03423	-.00124	.00392	.00022	.00101	-.00038
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10132.**	**61150.**	**3069.8**
Stddev	20.	142.	13.8
%RSD	.19324	.23287	.44985
#1	10144.	61269.	3085.6
#2	10109.	61188.	3063.7
#3	10143.	60992.	3060.0

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 20:14:26 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39742	10.037	.39212	.48448	.99926	.05007	9.9665
Stddev	.00078	.015	.00234	.00170	.00686	.00012	.0639
%RSD	.19741	.14886	.59686	.35046	.68696	.24571	.64080
#1	.39794	10.020	.39189	.48304	1.0007	.05009	10.032
#2	.39780	10.040	.39457	.48635	.99181	.05018	9.9043
#3	.39651	10.049	.38991	.48406	1.0053	.04994	9.9633
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04995	.19662	.49719	.50408	3.9345	48.425	1.0056
Stddev	.00017	.00028	.00293	.00085	.0512	.336	.0066
%RSD	.33202	.14096	.58868	.16772	1.3013	.69474	.65876
#1	.05012	.19642	.49505	.50367	3.9195	48.601	1.0066
#2	.04993	.19694	.50053	.50505	3.8924	48.037	.99858
#3	.04979	.19651	.49600	.50351	3.9915	48.637	1.0117
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.6539	.49874	1.0057	49.161	.50061	9.8144	.49889
Stddev	.1152	.00402	.0014	.263	.00048	.0158	.00214
%RSD	1.1928	.80691	.13850	.53437	.09607	.16108	.42961
#1	9.7007	.49641	1.0053	49.311	.50117	9.7964	.49662
#2	9.5228	.49643	1.0072	48.858	.50038	9.8207	.49919
#3	9.7383	.50339	1.0045	49.316	.50029	9.8261	.50087
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 20:14:26 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1787**	**.38972**	**5.0552**	**.98705**	**.99934**	**.99527**	**.50770**
Stddev	.0014	.00525	.0102	.00133	.00573	.00628	.00313
%RSD	.12146	1.3477	.20130	.13498	.57306	.63102	.61635
#1	1.1790	.38370	5.0434	.98688	1.0013	.99951	.50582
#2	1.1772	.39212	5.0610	.98581	.99288	.98805	.50598
#3	1.1800	.39334	5.0611	.98846	1.0038	.99824	.51132
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0147**	**.99253**	F **2.9496**
Stddev	.0017	.00087	.0127
%RSD	.16286	.08813	.43029
#1	1.0131	.99259	2.9390
#2	1.0147	.99336	2.9462
#3	1.0164	.99162	2.9637
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9868.5**	**58806.**	**2964.6**
Stddev	19.7	151.	1.8
%RSD	.19965	.25680	.06126
#1	9885.2	58672.	2966.0
#2	9873.5	58776.	2965.4
#3	9846.8	58969.	2962.6

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 20:18:03 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00132	-.00340	-.00049	.00031	.00192	.00008	-.02002	-.00008
Stddev	.00216	.00700	.00114	.00069	.00099	.00015	.01568	.00025
%RSD	163.88	205.77	232.07	221.43	51.768	179.89	78.343	317.94
#1	.00112	.00138	-.00179	.00077	.00240	-.00004	-.03265	.00011
#2	.00357	-.00014	-.00002	.00066	.00258	.00004	-.00247	-.00036
#3	-.00074	-.01144	.00034	-.00049	.00078	.00025	-.02493	.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00074	.00038	.00063	.02332	-.07555	-.00123	.02786	-.00147
Stddev	.00036	.00095	.00188	.02448	.09270	.00682	.09570	.00329
%RSD	47.976	247.76	298.28	104.94	122.70	553.52	343.47	223.66
#1	.00080	.00146	-.00061	.04476	-.06611	.00634	-.06528	.00195
#2	.00036	.00002	.00279	-.00335	-.17262	-.00687	.02294	-.00462
#3	.00107	-.00033	-.00029	.02856	.01207	-.00317	.12594	-.00174
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00377	.03770	.00015	.00241	-.00006	.00325	.00675	.03903
Stddev	.00019	.01915	.00084	.00431	.00100	.00269	.00806	.00132
%RSD	5.0202	50.799	576.30	178.78	1573.6	82.788	119.45	3.3700
#1	.00388	.01718	.00046	.00371	.00063	.00040	.01512	.04015
#2	.00389	.04082	.00078	-.00240	-.00121	.00574	.00608	.03936
#3	.00355	.05510	-.00080	.00592	.00039	.00360	-.00096	.03758
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 20:18:03 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00001**	**.00027**	**.00088**	**-.00106**	**.00004**	**-.00003**	**.00449**
Stddev	.00046	.00064	.00516	.00225	.00089	.00028	.00124
%RSD	6827.2	239.61	583.41	212.27	2439.5	892.40	27.555
#1	.00053	-.00043	-.00417	-.00308	-.00098	.00021	.00367
#2	-.00031	.00040	.00068	.00137	.00041	.00004	.00389
#3	-.00023	.00082	.00615	-.00147	.00068	-.00034	.00592
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9631.9**	**58363.**	**2969.7**
Stddev	4.5	93.	16.5
%RSD	.04685	.16010	.55454
#1	9627.4	58461.	2965.8
#2	9636.4	58275.	2987.8
#3	9631.8	58353.	2955.6

Approved: June 27, 2016

Sample Name: L1606098609 Acquired: 6/25/2016 20:22:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00081	.68625	.00096	.00637	.02063	.00019	116.68	.00011
Stddev	.00130	.00243	.00203	.00124	.00115	.00006	.34	.00020
%RSD	161.33	.35410	210.74	19.489	5.5807	30.863	.28955	179.93
#1	.00013	.68352	.00323	.00768	.02147	.00021	116.54	.00014
#2	.00231	.68708	-.00070	.00521	.01932	.00023	116.43	-.00010
#3	-.00002	.68816	.00037	.00620	.02109	.00012	117.06	.00029
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.00251	.00589	.80058	.71511	.00736	28.613	.02306
Stddev	.00014	.00088	.00100	.02799	.18771	.00368	.079	.00071
%RSD	23.262	34.910	17.057	3.4966	26.250	49.956	.27754	3.0949
#1	.00051	.00150	.00680	.80914	.91494	.01159	28.526	.02382
#2	.00075	.00293	.00481	.76931	.68792	.00556	28.632	.02293
#3	.00052	.00310	.00605	.82330	.54247	.00492	28.681	.02241
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00074	14.713	.00303	.03333	.00106	-.00272	.00365	4.8889
Stddev	.00024	.075	.00054	.00378	.00114	.00271	.00111	.0101
%RSD	32.746	.50774	17.920	11.335	108.20	99.808	30.510	.20564
#1	.00047	14.684	.00299	.03128	.00107	-.00577	.00444	4.8999
#2	.00095	14.656	.00359	.03769	-.00009	-.00056	.00238	4.8868
#3	.00081	14.797	.00251	.03102	.00220	-.00183	.00413	4.8801
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098609 Acquired: 6/25/2016 20:22:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00003**	**.51888**	**.00260**	**-.00246**	**.00196**	**.03426**	**.00729**
Stddev	.00032	.00275	.00453	.00099	.00037	.00011	.00028
%RSD	1222.6	.53076	174.15	40.180	18.856	.31556	3.8509
#1	.00009	.51844	-.00256	-.00245	.00220	.03417	.00723
#2	-.00032	.51637	.00591	-.00346	.00153	.03438	.00705
#3	.00031	.52183	.00446	-.00148	.00215	.03423	.00760
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9797.7**	**59515.**	**3093.6**
Stddev	11.0	17.	11.2
%RSD	.11267	.02773	.36079
#1	9791.0	59496.	3083.8
#2	9810.4	59520.	3105.7
#3	9791.6	59528.	3091.1

Approved: June 27, 2016

Sample Name: L1606098611 Acquired: 6/25/2016 20:25:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00025	.67808	-.00318	.08715	.06918	.00114	148.14	.00017
Stddev	.00037	.00569	.00382	.00198	.00070	.00012	.69	.00021
%RSD	151.84	.83972	120.10	2.2728	1.0151	10.946	.46623	125.05
#1	-.00014	.67215	-.00172	.08491	.06933	.00099	148.49	.00006
#2	.00061	.68350	-.00031	.08867	.06980	.00119	147.34	.00003
#3	.00027	.67861	-.00751	.08788	.06842	.00123	148.59	.00041
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00363	.00334	.07531	1.1555	1.9378	.02263	43.443	.16489
Stddev	.00025	.00064	.00057	.0192	.0114	.00045	.331	.00288
%RSD	6.9167	19.099	.75123	1.6613	.58882	1.9873	.76282	1.7487
#1	.00388	.00297	.07511	1.1584	1.9507	.02262	43.610	.16408
#2	.00338	.00297	.07595	1.1731	1.9292	.02219	43.061	.16250
#3	.00362	.00407	.07487	1.1351	1.9335	.02309	43.657	.16809
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00248	25.675	.00454	.04285	.00455	-.00035	-.00286	3.7826
Stddev	.00024	.125	.00067	.00695	.00113	.00106	.00555	.0053
%RSD	9.7863	.48812	14.843	16.217	24.766	306.62	194.07	.13999
#1	.00260	25.775	.00423	.04366	.00585	.00053	-.00927	3.7791
#2	.00220	25.535	.00409	.03554	.00387	-.00004	.00020	3.7887
#3	.00264	25.716	.00532	.04936	.00394	-.00153	.00049	3.7799
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098611 Acquired: 6/25/2016 20:25:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00008	2.0876	-.00638	-.00280	.00142	.06566	.00394
Stddev	.00055	.0096	.00925	.00066	.00022	.00017	.00141
%RSD	713.59	.45905	144.86	23.607	15.199	.25178	35.771
#1	-.00017	2.0923	-.00754	-.00222	.00163	.06585	.00254
#2	-.00031	2.0765	-.01500	-.00266	.00143	.06556	.00536
#3	.00071	2.0939	.00339	-.00351	.00120	.06556	.00393
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9695.5	58902.	3054.3
Stddev	10.1	156.	10.7
%RSD	.10398	.26470	.35100
#1	9691.1	59006.	3048.2
#2	9688.3	58978.	3066.6
#3	9707.0	58723.	3048.0

Approved: June 27, 2016

Sample Name: L1606098613 Acquired: 6/25/2016 20:29:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00174	5.5730	.00432	.03801	.11035	.00130	203.09	-.00001
Stddev	.00149	.0359	.00264	.00141	.00031	.00003	.17	.00016
%RSD	85.559	.64338	61.135	3.7127	.27868	2.0129	.08522	1090.1
#1	.00010	5.6007	.00735	.03819	.11034	.00128	203.27	.00012
#2	.00211	5.5858	.00310	.03932	.11006	.00133	203.05	.00003
#3	.00302	5.5325	.00251	.03651	.11067	.00131	202.93	-.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00373	.00967	.08024	7.4517	4.1196	.03087	79.593	.34313
Stddev	.00008	.00046	.00072	.0194	.0319	.00148	.239	.00219
%RSD	2.0959	4.7129	.90070	.26051	.77317	4.7986	.30050	.63957
#1	.00364	.00917	.07941	7.4293	4.0890	.03222	79.370	.34059
#2	.00377	.01007	.08072	7.4635	4.1171	.03112	79.565	.34436
#3	.00379	.00978	.08059	7.4624	4.1526	.02929	79.846	.34443
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00182	24.658	.00975	.17618	.00825	.00008	.00248	10.141
Stddev	.00040	.019	.00024	.00598	.00295	.00311	.00271	.073
%RSD	21.744	.07518	2.5015	3.3961	35.754	3886.1	109.35	.72041
#1	.00146	24.670	.00990	.17816	.00502	.00353	.00373	10.058
#2	.00175	24.669	.00947	.16946	.00891	-.00250	.00433	10.169
#3	.00224	24.637	.00989	.18093	.01080	-.00079	-.00063	10.196
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098613 Acquired: 6/25/2016 20:29:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00066	2.4743	.05397	-.00080	.01073	.41856	.00624
Stddev	.00027	.0052	.00665	.00236	.00131	.00229	.00115
%RSD	39.960	.21077	12.313	294.11	12.169	.54671	18.419
#1	.00089	2.4751	.05254	-.00353	.01217	.41596	.00756
#2	.00073	2.4791	.04816	.00063	.00962	.42027	.00548
#3	.00037	2.4687	.06122	.00049	.01040	.41945	.00567
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9619.4	58503.	3058.9
Stddev	38.3	135.	15.7
%RSD	.39844	.23102	.51315
#1	9576.0	58658.	3041.4
#2	9633.7	58414.	3071.7
#3	9648.5	58435.	3063.7

Approved: June 27, 2016

Sample Name: L1606098615 Acquired: 6/25/2016 20:33:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00201	.53689	.00109	.12769	.02932	.00091	261.33	.00014
Stddev	.00048	.01300	.00276	.00142	.00104	.00011	1.37	.00008
%RSD	23.740	2.4215	252.93	1.1120	3.5600	12.063	.52322	55.108
#1	.00148	.54532	.00299	.12707	.02978	.00081	261.04	.00021
#2	.00215	.54344	.00237	.12931	.03005	.00088	262.83	.00006
#3	.00240	.52192	-.00208	.12669	.02812	.00103	260.14	.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00075	.00167	.04808	1.7755	3.9837	.04482	96.121	.35461
Stddev	.00009	.00172	.00105	.0329	.0639	.00305	.330	.00305
%RSD	12.352	103.13	2.1885	1.8532	1.6044	6.8091	.34352	.86008
#1	.00068	.00125	.04714	1.8121	3.9806	.04242	96.172	.35269
#2	.00071	.00356	.04922	1.7483	4.0491	.04379	96.422	.35813
#3	.00085	.00020	.04788	1.7662	3.9214	.04825	95.768	.35302
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00080	71.586	.00152	.06528	.00543	-.00023	-.00042	6.3830
Stddev	.00049	.359	.00083	.00282	.00224	.00292	.00309	.0073
%RSD	60.493	.50193	55.060	4.3183	41.272	1275.7	729.65	.11354
#1	.00036	71.588	.00097	.06827	.00795	-.00316	.00167	6.3913
#2	.00132	71.945	.00248	.06488	.00468	-.00019	.00103	6.3782
#3	.00072	71.227	.00110	.06267	.00366	.00267	-.00398	6.3795
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098615 Acquired: 6/25/2016 20:33:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00066	6.9477	-.01868	-.00208	-.00036	.07416	.00363
Stddev	.00094	.0407	.00365	.00070	.00044	.00028	.00062
%RSD	142.12	.58617	19.553	33.577	122.09	.37861	17.066
#1	.00171	6.9428	-.02024	-.00215	-.00034	.07441	.00412
#2	.00039	6.9907	-.01451	-.00134	-.00081	.07386	.00293
#3	-.00011	6.9097	-.02130	-.00273	.00007	.07420	.00382
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9482.1	57521.	3041.1
Stddev	15.4	119.	13.0
%RSD	.16231	.20772	.42869
#1	9472.3	57649.	3053.5
#2	9499.8	57500.	3027.5
#3	9474.1	57412.	3042.2

Approved: June 27, 2016

Sample Name: L1606098617 Acquired: 6/25/2016 20:37:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00154	.20547	.00214	.04863	.05628	.00015	62.798	.00005
Stddev	.00260	.00707	.00399	.00222	.00076	.00003	.506	.00026
%RSD	168.84	3.4433	186.39	4.5608	1.3548	21.842	.80652	579.39
#1	.00162	.19836	-.00117	.04752	.05710	.00015	63.380	-.00016
#2	.00410	.20553	.00657	.05118	.05613	.00018	62.459	.00034
#3	-.00110	.21251	.00102	.04718	.05560	.00012	62.555	-.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00034	.00139	.00754	.49751	2.0420	.02056	20.601	.01571
Stddev	.00024	.00096	.00096	.02565	.1117	.00244	.034	.00089
%RSD	70.093	68.878	12.744	5.1552	5.4689	11.879	.16418	5.6341
#1	.00059	.00246	.00739	.47113	2.0254	.02185	20.628	.01672
#2	.00012	.00112	.00857	.52236	1.9395	.01774	20.613	.01507
#3	.00031	.00060	.00666	.49904	2.1610	.02209	20.563	.01535
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00090	26.598	.00080	.00703	.00246	-.00325	.00131	3.4906
Stddev	.00064	.178	.00058	.00500	.00246	.00605	.00554	.0173
%RSD	70.655	.66988	72.589	71.163	99.983	186.20	421.53	.49585
#1	.00155	26.804	.00039	.00280	.00327	-.00237	-.00267	3.5042
#2	.00088	26.495	.00146	.00574	-.00030	-.00969	.00764	3.4966
#3	.00027	26.495	.00054	.01256	.00441	.00231	-.00103	3.4711
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098617 Acquired: 6/25/2016 20:37:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00032	1.5150	-.00311	-.00236	-.00015	.01447	.00341
Stddev	.00072	.0088	.00641	.00068	.00005	.00011	.00060
%RSD	226.25	.58268	206.23	28.717	34.447	.73139	17.615
#1	.00066	1.5251	-.00263	-.00291	-.00009	.01435	.00287
#2	.00080	1.5088	.00305	-.00160	-.00015	.01454	.00406
#3	-.00051	1.5111	-.00975	-.00256	-.00020	.01452	.00331
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9830.7	59822.	3082.9
Stddev	15.5	147.	23.6
%RSD	.15717	.24641	.76693
#1	9813.6	59963.	3057.5
#2	9843.5	59835.	3104.4
#3	9835.1	59669.	3086.7

Approved: June 27, 2016

Sample Name: L1606098619 Acquired: 6/25/2016 20:41:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00113	.27749	.00069	.06930	.05323	.00039	10.214	.00034
Stddev	.00263	.00318	.00178	.00081	.00134	.00007	.065	.00038
%RSD	232.09	1.1459	258.13	1.1649	2.5186	17.965	.63738	111.28
#1	.00355	.27762	.00105	.06852	.05477	.00042	10.240	.00066
#2	-.00167	.28061	.00227	.06924	.05251	.00044	10.140	.00044
#3	.00153	.27425	-.00125	.07014	.05240	.00031	10.262	-.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00108	.00071	.02585	.24748	1.2811	.01465	4.6162	.00999
Stddev	.00051	.00114	.00067	.00474	.0573	.00378	.0318	.00359
%RSD	46.988	159.40	2.5887	1.9167	4.4759	25.817	.68816	35.904
#1	.00062	.00202	.02645	.25048	1.2371	.01153	4.5886	.01278
#2	.00162	-.00003	.02513	.24201	1.3460	.01356	4.6509	.01123
#3	.00099	.00015	.02599	.24994	1.2604	.01885	4.6090	.00594
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00123	228.52	.00044	.07963	.00371	-.00207	.00619	2.5103
Stddev	.00028	1.82	.00063	.00808	.00266	.00077	.00393	.0173
%RSD	22.669	.79720	141.35	10.151	71.562	37.379	63.490	.68709
#1	.00153	229.52	.00003	.07034	.00201	-.00126	.00175	2.5018
#2	.00118	226.42	.00013	.08504	.00235	-.00280	.00758	2.4989
#3	.00097	229.62	.00117	.08351	.00677	-.00216	.00924	2.5301
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098619 Acquired: 6/25/2016 20:41:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00007	.22967	.00683	-.00264	.00219	.02256	.00408
Stddev	.00091	.00074	.00628	.00047	.00030	.00021	.00266
%RSD	1371.8	.32350	91.943	17.727	13.766	.92425	65.101
#1	-.00024	.23041	.00516	-.00255	.00222	.02259	.00112
#2	-.00065	.22892	.00155	-.00314	.00188	.02235	.00625
#3	.00109	.22967	.01377	-.00222	.00248	.02276	.00487
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9702.8	58480.	3067.3
Stddev	45.7	28.	15.6
%RSD	.47062	.04775	.50880
#1	9725.9	58510.	3053.5
#2	9732.4	58474.	3084.2
#3	9650.3	58455.	3064.3

Approved: June 27, 2016

Sample Name: L1606098621 Acquired: 6/25/2016 20:45:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00057	.00121	.00179	.05746	.02986	.00029	128.59	-.00000
Stddev	.00056	.00275	.00384	.00358	.00132	.00008	.53	.00025
%RSD	97.573	226.34	214.60	6.2318	4.4201	27.216	.41011	6271.6
#1	.00040	-.00100	.00602	.06157	.03100	.00034	129.12	-.00029
#2	.00120	.00036	-.00148	.05576	.03017	.00020	128.07	.00019
#3	.00012	.00429	.00083	.05504	.02841	.00032	128.57	.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00019	.00173	.01434	.05981	2.5098	.02550	33.269	.15420
Stddev	.00017	.00097	.00185	.00891	.0950	.00146	.060	.00170
%RSD	92.832	55.806	12.905	14.892	3.7837	5.7272	.18077	1.1012
#1	.00010	.00269	.01260	.05052	2.5555	.02390	33.334	.15465
#2	.00007	.00177	.01629	.06064	2.4007	.02586	33.216	.15563
#3	.00038	.00075	.01412	.06828	2.5733	.02675	33.257	.15232
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00056	47.893	-.00000	.01004	.00162	-.00276	-.00417	3.5790
Stddev	.00060	.257	.00189	.00801	.00056	.00235	.00172	.0047
%RSD	106.13	.53712	269880.	79.763	34.469	85.173	41.117	.13102
#1	.00071	48.149	.00182	.00080	.00218	-.00406	-.00514	3.5827
#2	-.00009	47.634	.00013	.01426	.00163	-.00417	-.00519	3.5737
#3	.00107	47.897	-.00195	.01505	.00106	-.00005	-.00219	3.5806
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098621 Acquired: 6/25/2016 20:45:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00024**	**1.8050**	**-.02231**	**-.00291**	**.00001**	**.01004**	**.00207**
Stddev	.00095	.0120	.00564	.00294	.00125	.00002	.00101
%RSD	393.73	.66689	25.288	100.88	21161.	.16952	48.549
#1	.00057	1.8173	-.02371	-.00622	-.00119	.01005	.00218
#2	-.00128	1.7932	-.02711	-.00189	.00130	.01004	.00102
#3	-.00001	1.8046	-.01610	-.00062	-.00009	.01002	.00302
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9662.5**	**58180.**	**3042.4**
Stddev	14.0	47.	7.3
%RSD	.14521	.08141	.24078
#1	9658.0	58146.	3035.1
#2	9678.2	58161.	3042.4
#3	9651.2	58234.	3049.8

Approved: June 27, 2016

Sample Name: L1606098623 Acquired: 6/25/2016 20:49:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00139	.10705	.00198	.41392	.02623	.00027	243.17	.00017
Stddev	.00218	.00541	.00096	.00428	.00052	.00003	1.41	.00011
%RSD	157.63	5.0515	48.506	1.0332	1.9773	12.148	.58072	65.505
#1	.00177	.11044	.00295	.41225	.02682	.00026	243.56	.00005
#2	.00336	.10081	.00102	.41878	.02603	.00030	244.34	.00020
#3	-.00096	.10989	.00198	.41074	.02584	.00024	241.60	.00026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00008	.00125	.02256	.13395	6.1008	.01177	68.965	.45308
Stddev	.00021	.00090	.00053	.00942	.0335	.00473	.254	.00553
%RSD	267.67	72.031	2.3661	7.0298	.54942	40.196	.36871	1.2196
#1	-.00012	.00056	.02318	.12309	6.1028	.01598	68.820	.44996
#2	.00031	.00093	.02221	.13890	6.0663	.00665	69.259	.45946
#3	.00005	.00227	.02231	.13986	6.1333	.01267	68.816	.44983
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00044	66.667	.00075	.01360	.00015	-.00234	.00238	3.6345
Stddev	.00039	.343	.00118	.00476	.00283	.00395	.00376	.0082
%RSD	88.357	.51401	157.96	35.012	1938.3	168.93	158.02	.22548
#1	.00001	66.765	.00108	.01427	-.00296	-.00415	.00669	3.6345
#2	.00056	66.950	-.00056	.00853	.00082	.00219	.00057	3.6426
#3	.00076	66.286	.00172	.01798	.00259	-.00506	-.00014	3.6263
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098623 Acquired: 6/25/2016 20:49:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	1.6999	-.02956	-.00156	.00030	.00857	.00329
Stddev	.00028	.0081	.00507	.00230	.00048	.00011	.00135
%RSD	598.12	.47536	17.163	147.79	160.42	1.3327	41.067
#1	.00005	1.7017	-.03255	.00072	-.00015	.00844	.00173
#2	.00033	1.7069	-.02370	-.00388	.00023	.00859	.00400
#3	-.00023	1.6911	-.03242	-.00151	.00081	.00866	.00415
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9416.4	57429.	3035.8
Stddev	48.0	130.	25.1
%RSD	.51009	.22575	.82551
#1	9467.2	57482.	3016.2
#2	9410.3	57282.	3027.1
#3	9371.8	57525.	3064.1

Approved: June 27, 2016

Sample Name: L1606098625 Acquired: 6/25/2016 20:52:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00225	2.3612	.00034	.04147	.11975	.00023	121.59	.00037
Stddev	.00036	.0069	.00441	.00209	.00117	.00012	.33	.00010
%RSD	15.843	.29306	1297.6	5.0487	.97876	50.808	.26806	27.630
#1	.00211	2.3674	.00539	.03961	.11950	.00036	121.76	.00028
#2	.00199	2.3537	-.00273	.04106	.11872	.00014	121.21	.00033
#3	.00266	2.3625	-.00164	.04374	.12102	.00019	121.79	.00048
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00161	.00414	.00666	2.3953	2.5999	.01491	37.526	1.0679
Stddev	.00025	.00072	.00102	.0202	.0477	.00333	.332	.0124
%RSD	15.696	17.478	15.382	.84513	1.8346	22.308	.88561	1.1598
#1	.00188	.00385	.00549	2.3789	2.6109	.01808	37.493	1.0637
#2	.00138	.00496	.00714	2.4179	2.5476	.01145	37.212	1.0581
#3	.00156	.00361	.00736	2.3890	2.6411	.01519	37.874	1.0818
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00064	18.226	.00269	.05391	.00651	-.00190	.00505	5.9129
Stddev	.00020	.052	.00077	.00146	.00208	.00291	.00127	.0400
%RSD	30.881	.28797	28.786	2.7065	31.918	152.94	25.077	.67624
#1	.00067	18.245	.00228	.05412	.00658	-.00209	.00366	5.9121
#2	.00082	18.167	.00221	.05526	.00440	.00110	.00533	5.9533
#3	.00043	18.267	.00358	.05236	.00855	-.00472	.00614	5.8733
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098625 Acquired: 6/25/2016 20:52:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00035**	**.99462**	**.03044**	**-.00185**	**.00457**	**.15726**	**.00437**
Stddev	.00053	.00442	.00913	.00019	.00047	.00042	.00072
%RSD	150.91	.44485	29.990	10.187	10.282	.27017	16.565
#1	.00010	.99601	.03923	-.00207	.00503	.15763	.00393
#2	-.00093	.98967	.02100	-.00171	.00410	.15735	.00520
#3	-.00022	.99819	.03111	-.00178	.00457	.15680	.00397
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9574.3**	**58285.**	**3022.7**
Stddev	18.0	80.	21.3
%RSD	.18820	.13801	.70548
#1	9553.8	58363.	3008.3
#2	9581.8	58202.	3047.2
#3	9587.4	58291.	3012.6

Approved: June 27, 2016

Sample Name: L1606098627 Acquired: 6/25/2016 20:56:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00254	4.6384	.00407	.03588	.08487	.00132	200.73	.00005
Stddev	.00192	.0173	.00097	.00149	.00032	.00011	.32	.00023
%RSD	75.499	.37394	23.741	4.1641	.37451	8.4866	.15998	422.09
#1	.00186	4.6215	.00464	.03631	.08521	.00120	200.37	.00006
#2	.00106	4.6377	.00461	.03711	.08483	.00141	200.85	-.00018
#3	.00471	4.6561	.00295	.03422	.08457	.00136	200.98	.00028
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00295	.00847	.07661	6.3761	4.0467	.02688	78.393	.28594
Stddev	.00029	.00096	.00098	.0153	.0946	.00177	.290	.00214
%RSD	9.9470	11.337	1.2842	.24023	2.3374	6.5906	.37011	.74836
#1	.00263	.00958	.07761	6.3586	3.9534	.02860	78.140	.28832
#2	.00299	.00791	.07658	6.3825	4.0443	.02506	78.709	.28419
#3	.00322	.00791	.07564	6.3872	4.1425	.02698	78.329	.28530
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00106	25.068	.00713	.15173	.00678	-.00194	.00054	8.9965
Stddev	.00058	.104	.00038	.00131	.00344	.00137	.00531	.0286
%RSD	55.068	.41293	5.2660	.86550	50.818	70.752	976.81	.31838
#1	.00130	24.949	.00680	.15078	.01021	-.00340	-.00348	9.0111
#2	.00040	25.140	.00754	.15323	.00679	-.00068	-.00144	9.0150
#3	.00149	25.114	.00704	.15117	.00333	-.00174	.00656	8.9635
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098627 Acquired: 6/25/2016 20:56:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00089	2.4772	.04490	-.00127	.00901	.16031	.00422
Stddev	.00154	.0067	.00796	.00182	.00061	.00117	.00055
%RSD	173.10	.26965	17.717	142.95	6.7328	.72957	12.944
#1	.00142	2.4695	.04447	-.00241	.00911	.16067	.00359
#2	.00210	2.4809	.03717	.00082	.00836	.16126	.00459
#3	-.00085	2.4812	.05306	-.00223	.00956	.15900	.00448
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9465.6	57841.	2995.1
Stddev	68.7	160.	17.6
%RSD	.72586	.27642	.58883
#1	9445.1	57701.	3013.7
#2	9409.5	58015.	2978.6
#3	9542.2	57807.	2993.1

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 21:00:36 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40235**	**10.185**	**.39842**	**.49195**	**1.0203**	**.05034**	**10.153**
Stddev	.00325	.028	.00193	.00194	.0055	.00031	.030
%RSD	.80675	.27855	.48505	.39452	.53912	.62290	.29151
#1	.39985	10.214	.39899	.49235	1.0140	.05068	10.126
#2	.40602	10.183	.40000	.49365	1.0227	.05026	10.185
#3	.40118	10.157	.39627	.48984	1.0242	.05007	10.150
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05123**	**.19988**	**.50288**	**.51330**	**4.0070**	**49.596**	**1.0284**
Stddev	.00026	.00047	.00335	.00193	.0033	.230	.0041
%RSD	.51360	.23712	.66679	.37527	.08271	.46359	.40302
#1	.05102	.20030	.50672	.51551	4.0072	49.649	1.0258
#2	.05153	.19997	.50051	.51234	4.0102	49.344	1.0261
#3	.05115	.19937	.50141	.51204	4.0036	49.795	1.0332
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.5669**	**.50781**	**1.0249**	**50.362**	**.50842**	**9.9814**	**.50557**
Stddev	.0485	.00742	.0065	.234	.00100	.0181	.00207
%RSD	.50657	1.4609	.62973	.46520	.19680	.18135	.40980
#1	9.6226	.50131	1.0311	50.266	.50877	9.9736	.50596
#2	9.5340	.50622	1.0255	50.191	.50920	10.002	.50742
#3	9.5442	.51589	1.0183	50.629	.50729	9.9685	.50333
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 21:00:36 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2045**	**.39202**	**5.1444**	**1.0010**	**1.0190**	**1.0190**	**.51206**
Stddev	.0033	.00153	.0072	.0031	.0045	.0084	.00251
%RSD	.27464	.38961	.14055	.30711	.44063	.82168	.49020
#1	1.2010	.39059	5.1361	1.0039	1.0167	1.0152	.51490
#2	1.2047	.39363	5.1492	1.0015	1.0162	1.0132	.51015
#3	1.2076	.39184	5.1479	.99778	1.0242	1.0286	.51112
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0223**	**1.0141**	F **2.9929**
Stddev	.0031	.0001	.0241
%RSD	.30006	.01208	.80650
#1	1.0229	1.0141	3.0133
#2	1.0190	1.0142	2.9663
#3	1.0251	1.0139	2.9992
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9649.2**	**58649.**	**2955.6**
Stddev	2.7	54.	16.9
%RSD	.02842	.09258	.57071
#1	9646.1	58667.	2944.4
#2	9650.3	58588.	2975.0
#3	9651.2	58692.	2947.5

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 21:04:14 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00098**	**-.00081**	**.00091**	**.00315**	**.00292**	**.00009**	**-.03474**	**.00002**
Stddev	.00124	.00864	.00092	.00201	.00133	.00005	.02919	.00008
%RSD	125.96	1064.0	100.48	64.019	45.548	52.705	84.022	417.72
#1	.00043	-.00670	.00067	.00330	.00336	.00005	-.02620	.00002
#2	-.00151	.00910	.00014	.00508	.00142	.00009	-.01077	-.00006
#3	-.00188	-.00483	.00193	.00106	.00397	.00014	-.06724	.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00004**	**.00125**	**.00083**	**.02088**	**-.14334**	**.00212**	**-.02955**	**.00114**
Stddev	.00066	.00051	.00122	.01391	.04593	.00105	.09306	.00075
%RSD	1530.9	41.168	147.43	66.595	32.046	49.608	314.87	66.235
#1	.00067	.00069	.00144	.03293	-.09030	.00098	-.05631	.00047
#2	-.00016	.00170	-.00058	.02406	-.17011	.00232	-.10630	.00196
#3	-.00064	.00135	.00162	.00566	-.16960	.00305	.07395	.00100
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00323**	**.05080**	**-.00020**	**.00483**	**.00058**	**.00191**	**.00111**	**.03921**
Stddev	.00021	.02575	.00089	.00353	.00235	.00387	.00242	.00153
%RSD	6.4882	50.676	441.28	73.104	406.34	202.21	216.78	3.9059
#1	.00300	.07783	.00045	.00259	-.00201	.00626	-.00135	.03746
#2	.00339	.02657	-.00122	.00300	.00257	.00062	.00121	.03991
#3	.00331	.04802	.00016	.00889	.00117	-.00114	.00348	.04027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 21:04:14 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	.00029	-.00007	-.00513	.00041	-.00001	.00470
Stddev	.00027	.00000	.00626	.00369	.00015	.00012	.00055
%RSD	55.343	1.4693	8856.4	71.949	37.134	916.05	11.589
#1	.00081	.00028	.00338	-.00666	.00031	.00010	.00427
#2	.00033	.00029	-.00730	-.00092	.00059	.00000	.00453
#3	.00034	.00028	.00370	-.00780	.00035	-.00014	.00531
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9506.6	58497.	2929.7
Stddev	22.3	113.	12.0
%RSD	.23433	.19372	.40841
#1	9532.3	58628.	2919.5
#2	9494.8	58435.	2926.6
#3	9492.6	58429.	2942.8

Approved: June 27, 2016

Sample Name: L1606098629 Acquired: 6/25/2016 21:08:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00070	1.3154	-.00179	.12516	.05800	.00081	247.18	.00018
Stddev	.00097	.0064	.00105	.00195	.00108	.00003	.31	.00008
%RSD	137.44	.48564	58.530	1.5550	1.8687	3.6829	.12350	43.911
#1	.00160	1.3100	-.00292	.12302	.05810	.00083	246.92	.00018
#2	.00082	1.3224	-.00085	.12682	.05904	.00078	247.52	.00010
#3	-.00032	1.3138	-.00160	.12563	.05687	.00083	247.09	.00026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00119	.00401	.04528	2.4364	3.9714	.04409	89.669	.34274
Stddev	.00048	.00063	.00072	.0178	.1152	.00255	.458	.00095
%RSD	40.012	15.569	1.5917	.73079	2.9014	5.7752	.51098	.27642
#1	.00152	.00449	.04490	2.4275	3.9851	.04702	90.007	.34368
#2	.00141	.00424	.04612	2.4569	4.0792	.04281	89.852	.34276
#3	.00065	.00331	.04484	2.4248	3.8500	.04243	89.147	.34178
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00084	72.159	.00306	.12842	.00472	-.00245	-.00232	7.1068
Stddev	.00019	.103	.00116	.00307	.00202	.00283	.00582	.0122
%RSD	22.591	.14315	38.091	2.3898	42.715	115.64	250.20	.17182
#1	.00062	72.124	.00297	.12975	.00386	-.00541	-.00257	7.1178
#2	.00090	72.275	.00194	.13059	.00327	-.00216	-.00801	7.1089
#3	.00099	72.077	.00426	.12491	.00702	.00023	.00361	7.0937
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098629 Acquired: 6/25/2016 21:08:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00063**	**6.8039**	**.00261**	**-.00311**	**.00203**	**.14656**	**.00713**
Stddev	.00109	.0140	.00201	.00153	.00039	.00030	.00139
%RSD	174.70	.20570	77.238	49.135	18.962	.20399	19.505
#1	.00026	6.7979	.00196	-.00298	.00228	.14634	.00715
#2	-.00024	6.8199	.00487	-.00165	.00159	.14690	.00573
#3	.00185	6.7939	.00100	-.00469	.00223	.14643	.00851
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9470.9**	**57988.**	**3007.5**
Stddev	19.0	71.	5.1
%RSD	.20079	.12320	.16890
#1	9472.3	57910.	3001.8
#2	9451.2	58005.	3011.5
#3	9489.2	58049.	3009.1

Approved: June 27, 2016

Sample Name: L1606098631S Acquired: 6/25/2016 21:12:02 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.18827	4.8059	.19457	.97775	.51082	.02368	136.28	.02401
Stddev	.00146	.0060	.00214	.00486	.00158	.00008	.55	.00011
%RSD	.77669	.12402	1.0980	.49690	.30853	.31910	.40245	.44044
#1	.18949	4.8104	.19700	.98276	.50954	.02372	135.66	.02402
#2	.18665	4.7991	.19298	.97305	.51258	.02359	136.67	.02389
#3	.18867	4.8083	.19374	.97744	.51033	.02373	136.53	.02410
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09298	.23341	.25035	1.9823	26.282	.52091	38.371	.48555
Stddev	.00063	.00083	.00168	.0210	.198	.00132	.105	.00060
%RSD	.67753	.35580	.67127	1.0601	.75241	.25420	.27268	.12393
#1	.09358	.23411	.25193	1.9586	26.065	.52066	38.253	.48489
#2	.09232	.23249	.25054	1.9895	26.453	.52234	38.408	.48570
#3	.09303	.23364	.24858	1.9987	26.329	.51972	38.452	.48607
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.49539	77.049	.23394	4.7766	.23491	.57914	.18644	6.1257
Stddev	.00233	.345	.00087	.0076	.00312	.00333	.00623	.0072
%RSD	.47123	.44745	.37273	.15954	1.3288	.57493	3.3395	.11822
#1	.49753	76.664	.23461	4.7848	.23800	.57665	.18760	6.1339
#2	.49574	77.327	.23426	4.7754	.23176	.57786	.17971	6.1201
#3	.49290	77.158	.23296	4.7697	.23495	.58292	.19200	6.1230
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098631S Acquired: 6/25/2016 21:12:02 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47308**	**2.3488**	**.46537**	**.23713**	**.48550**	**.47406**	**.01603**
Stddev	.00180	.0057	.00826	.00130	.00209	.00072	.00324
%RSD	.38025	.24432	1.7741	.54866	.42967	.15151	20.234
#1	.47478	2.3426	.46034	.23744	.48500	.47482	.01233
#2	.47325	2.3539	.46086	.23825	.48371	.47397	.01737
#3	.47120	2.3498	.47490	.23571	.48779	.47340	.01839
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9509.0**	**58216.**	**3011.6**
Stddev	7.0	203.	15.7
%RSD	.07407	.34863	.52135
#1	9501.0	58067.	3028.6
#2	9511.6	58447.	2997.6
#3	9514.3	58134.	3008.7

Approved: June 27, 2016

Sample Name: L1606098633SD Acquired: 6/25/2016 21:15:41 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-06

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19322**	**4.9340**	**.19821**	**1.0094**	**.52302**	**.02438**	**139.13**	**.02459**
Stddev	.00378	.0145	.00351	.0014	.00226	.00009	.67	.00034
%RSD	1.9571	.29336	1.7708	.13572	.43296	.36107	.48341	1.3705
#1	.18918	4.9313	.20201	1.0102	.52283	.02442	139.33	.02427
#2	.19667	4.9211	.19752	1.0078	.52537	.02443	139.68	.02494
#3	.19382	4.9497	.19509	1.0102	.52086	.02428	138.38	.02455
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09514**	**.23910**	**.25336**	**2.0284**	**26.984**	**.53592**	**39.198**	**.43704**
Stddev	.00042	.00077	.00150	.0114	.096	.00060	.184	.00451
%RSD	.44105	.32091	.59054	.56072	.35526	.11184	.46916	1.0308
#1	.09541	.23859	.25475	2.0343	27.015	.53581	38.996	.43232
#2	.09535	.23873	.25356	2.0357	27.059	.53539	39.242	.44129
#3	.09465	.23999	.25178	2.0153	26.876	.53657	39.356	.43751
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50931**	**78.441**	**.23759**	**4.8949**	**.24046**	**.58794**	**.19015**	**6.2655**
Stddev	.00186	.336	.00142	.0178	.00324	.00596	.00110	.0092
%RSD	.36520	.42898	.59954	.36432	1.3470	1.0133	.57938	.14668
#1	.51115	78.549	.23885	4.9125	.24157	.58732	.19136	6.2719
#2	.50933	78.710	.23604	4.8953	.23682	.59419	.18990	6.2695
#3	.50743	78.063	.23787	4.8768	.24301	.58232	.18920	6.2549
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098633SD Acquired: 6/25/2016 21:15:41 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-06

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48269**	**2.3976**	**.47749**	**.23961**	**.50144**	**.48370**	**.00655**
Stddev	.00278	.0083	.00742	.00256	.00242	.00154	.00139
%RSD	.57608	.34786	1.5549	1.0679	.48318	.31822	21.270
#1	.48505	2.3997	.47422	.24207	.50191	.48531	.00572
#2	.48340	2.4047	.47226	.23981	.49882	.48355	.00816
#3	.47963	2.3884	.48599	.23696	.50360	.48224	.00577
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9541.3**	**57975.**	**3017.9**
Stddev	6.7	58.	25.7
%RSD	.06977	.09979	.85022
#1	9537.4	57946.	3002.1
#2	9549.0	58042.	3004.1
#3	9537.6	57938.	3047.5

Approved: June 27, 2016

Sample Name: L1606098635S Acquired: 6/25/2016 21:19:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19277**	**4.9885**	**.19579**	**1.3512**	**.50932**	**.02402**	**246.59**	**.02462**
Stddev	.00213	.0073	.00137	.0073	.00284	.00001	1.33	.00040
%RSD	1.1047	.14711	.70076	.54069	.55753	.03190	.53842	1.6295
#1	.19260	4.9802	.19687	1.3428	.50650	.02402	245.12	.02488
#2	.19498	4.9915	.19425	1.3565	.50929	.02401	246.97	.02416
#3	.19074	4.9939	.19626	1.3542	.51218	.02403	247.69	.02483
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09311**	**.23510**	**.26651**	**1.9710**	**30.025**	**.51865**	**72.396**	**.67936**
Stddev	.00111	.00066	.00283	.0124	.152	.00282	.624	.00361
%RSD	1.1973	.28134	1.0623	.63080	.50639	.54413	.86161	.53080
#1	.09375	.23474	.26754	1.9611	29.892	.51908	71.843	.67568
#2	.09182	.23470	.26331	1.9669	29.993	.51564	72.272	.67951
#3	.09375	.23586	.26869	1.9849	30.191	.52123	73.072	.68289
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49951**	**91.017**	**.23206**	**4.8553**	**.23760**	**.58301**	**.18517**	**6.5213**
Stddev	.00475	.491	.00241	.0437	.00246	.00711	.00329	.0728
%RSD	.95179	.53976	1.0367	.89938	1.0354	1.2190	1.7758	1.1160
#1	.50253	90.509	.23402	4.8765	.23815	.58650	.18639	6.5577
#2	.49403	91.051	.22938	4.8051	.23491	.57483	.18144	6.4375
#3	.50197	91.490	.23278	4.8843	.23974	.58769	.18767	6.5687
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098635S Acquired: 6/25/2016 21:19:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-07

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47340**	**2.1822**	**.45682**	**.23569**	**.48957**	**.47212**	**.00440**
Stddev	.00586	.0116	.00441	.00472	.00317	.00386	.00078
%RSD	1.2388	.53009	.96525	2.0030	.64754	.81664	17.679
#1	.47626	2.1692	.45177	.23827	.48595	.47646	.00457
#2	.46666	2.1856	.45986	.23024	.49090	.46909	.00508
#3	.47729	2.1916	.45885	.23856	.49186	.47080	.00355
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9474.0**	**57681.**	**2990.6**
Stddev	86.2	193.	10.0
%RSD	.90991	.33483	.33371
#1	9421.7	57902.	3000.6
#2	9573.5	57545.	2990.5
#3	9426.8	57595.	2980.6

Approved: June 27, 2016

Sample Name: L1606098637SD Acquired: 6/25/2016 21:22:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.18854**	**4.9112**	**.19745**	**1.3462**	**.50578**	**.02318**	**246.84**	**.02457**
Stddev	.00237	.0122	.00114	.0020	.00229	.00008	1.01	.00037
%RSD	1.2581	.24857	.57516	.14562	.45291	.33396	.40736	1.5133
#1	.18824	4.9085	.19618	1.3456	.50836	.02311	247.86	.02498
#2	.18633	4.9245	.19780	1.3447	.50496	.02326	245.85	.02445
#3	.19105	4.9006	.19837	1.3484	.50401	.02315	246.80	.02427
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09368**	**.23124**	**.24645**	**1.9652**	**29.866**	**.51243**	**71.961**	**.67062**
Stddev	.00032	.00174	.00162	.0103	.162	.00260	.245	.00120
%RSD	.34157	.75435	.65887	.52496	.54204	.50740	.33982	.17871
#1	.09405	.23044	.24793	1.9540	30.010	.51110	72.114	.67080
#2	.09349	.23325	.24471	1.9744	29.691	.51543	71.679	.66935
#3	.09350	.23005	.24671	1.9671	29.897	.51077	72.089	.67172
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50362**	**90.574**	**.23429**	**4.8998**	**.23803**	**.58838**	**.19013**	**6.6122**
Stddev	.00184	.311	.00137	.0310	.00392	.00166	.00571	.0196
%RSD	.36548	.34377	.58644	.63360	1.6481	.28131	3.0018	.29638
#1	.50535	90.891	.23522	4.9314	.24008	.58995	.18521	6.6237
#2	.50169	90.269	.23272	4.8693	.24051	.58666	.19639	6.5896
#3	.50382	90.562	.23494	4.8987	.23351	.58854	.18878	6.6233
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098637SD Acquired: 6/25/2016 21:22:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-08

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47518**	**2.1930**	**.45556**	**.23535**	**.47970**	**.47047**	**.00421**
Stddev	.00224	.0050	.00619	.00164	.00083	.00231	.00126
%RSD	.47229	.22774	1.3595	.69822	.17220	.49043	29.868
#1	.47661	2.1987	.46015	.23423	.48048	.47206	.00301
#2	.47259	2.1891	.45801	.23458	.47883	.46782	.00552
#3	.47634	2.1914	.44851	.23723	.47980	.47152	.00409
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9517.7**	**58425.**	**3034.0**
Stddev	30.6	371.	4.3
%RSD	.32191	.63434	.14141
#1	9523.8	58332.	3029.9
#2	9544.8	58110.	3038.5
#3	9484.5	58834.	3033.6

Approved: June 27, 2016

Sample Name: L1606098639 Acquired: 6/25/2016 21:26:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00004**	**.00422**	**.00040**	**.00975**	**.00108**	**.00009**	**.14419**	**-.00005**
Stddev	.00042	.00758	.00073	.00126	.00087	.00008	.01744	.00009
%RSD	1118.0	179.89	181.74	12.897	80.476	90.859	12.096	178.96
#1	.00028	.01060	-.00028	.01106	.00209	.00013	.16280	-.00014
#2	.00012	-.00417	.00116	.00964	.00057	-.00000	.14155	-.00004
#3	-.00051	.00622	.00032	.00856	.00059	.00013	.12822	.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00044**	**.00035**	**.00305**	**.04079**	**-.00706**	**.00203**	**-.02733**	**-.00024**
Stddev	.00042	.00030	.00090	.02484	.04162	.00092	.03833	.00264
%RSD	93.882	85.053	29.494	60.907	589.18	45.533	140.24	1117.6
#1	.00013	.00037	.00205	.05000	-.05467	.00295	-.00835	-.00277
#2	.00091	.00004	.00332	.05971	.01106	.00204	-.00220	-.00044
#3	.00028	.00063	.00378	.01266	.02242	.00110	-.07145	.00250
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00046**	**.23384**	**.00002**	**.00301**	**.00223**	**-.00336**	**.00424**	**.10834**
Stddev	.00025	.02078	.00057	.00488	.00185	.00118	.00213	.00216
%RSD	54.203	8.8874	3135.0	161.94	83.035	35.176	50.129	1.9925
#1	.00023	.25768	.00001	-.00120	.00258	-.00441	.00208	.11078
#2	.00073	.22428	.00059	.00188	.00389	-.00208	.00633	.10755
#3	.00043	.21955	-.00055	.00835	.00023	-.00360	.00432	.10668
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098639 Acquired: 6/25/2016 21:26:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00036	.00188	.00201	-.00117	.00072	.00941	.00130
Stddev	.00073	.00051	.00315	.00081	.00019	.00001	.00060
%RSD	203.49	27.359	156.60	69.426	26.854	.12705	46.035
#1	.00061	.00146	.00564	-.00068	.00081	.00940	.00116
#2	.00093	.00173	.00038	-.00211	.00085	.00941	.00078
#3	-.00047	.00245	.00001	-.00072	.00050	.00942	.00196
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9979.7	61621.	3102.9
Stddev	10.2	118.	6.8
%RSD	.10173	.19116	.21926
#1	9983.1	61738.	3101.0
#2	9987.7	61621.	3110.5
#3	9968.3	61503.	3097.3

Approved: June 27, 2016

Sample Name: L1606098641 Acquired: 6/25/2016 21:30:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00155	.00228	-.00074	.00681	.00240	.00007	2.3285	.00015
Stddev	.00195	.00437	.00132	.00259	.00010	.00002	.0225	.00014
%RSD	125.93	191.41	178.17	38.016	4.0011	29.007	.96410	94.650
#1	.00016	.00624	-.00146	.00963	.00251	.00005	2.3542	.00003
#2	.00071	.00302	-.00156	.00453	.00235	.00009	2.3184	.00011
#3	.00378	-.00241	.00078	.00628	.00234	.00008	2.3128	.00030
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00006	.00014	.00163	.03413	.24665	.00092	.81266	-.00079
Stddev	.00068	.00062	.00075	.01156	.02854	.00423	.01745	.00318
%RSD	1106.1	456.64	45.829	33.860	11.573	458.16	2.1475	404.35
#1	-.00033	.00085	.00092	.03095	.24915	.00580	.81529	.00231
#2	.00085	-.00023	.00241	.02450	.27386	-.00166	.79404	-.00404
#3	-.00034	-.00022	.00156	.04694	.21693	-.00138	.82864	-.00063
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	1.5234	-.00034	.00759	.00193	-.00110	-.00056	1.9225
Stddev	.00028	.0181	.00053	.00504	.00328	.00190	.00524	.0055
%RSD	316.47	1.1895	155.07	66.429	170.02	173.25	938.70	.28386
#1	-.00021	1.5412	-.00096	.00263	.00004	-.00329	-.00638	1.9218
#2	.00035	1.5240	.00002	.00743	.00571	-.00014	.00092	1.9283
#3	.00013	1.5050	-.00009	.01271	.00003	.00014	.00379	1.9175
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098641 Acquired: 6/25/2016 21:30:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00003**	**.02979**	**.00925**	**-.00425**	**.00079**	**.00575**	**.00161**
Stddev	.00070	.00033	.00260	.00161	.00118	.00006	.00128
%RSD	2613.1	1.1094	28.082	37.920	148.25	1.0430	79.495
#1	-.00014	.02957	.00917	-.00541	.00147	.00579	.00219
#2	.00072	.03017	.01188	-.00492	.00148	.00578	.00014
#3	-.00067	.02963	.00669	-.00241	-.00057	.00568	.00250
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9984.0**	**61505.**	**3103.5**
Stddev	25.4	97.	8.6
%RSD	.25425	.15693	.27670
#1	10013.	61422.	3093.7
#2	9966.4	61611.	3106.9
#3	9972.6	61483.	3109.8

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 21:34:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40382**	**10.228**	**.39903**	**.49252**	**1.0184**	**.05057**	**10.132**
Stddev	.00317	.045	.00339	.00303	.0029	.00010	.084
%RSD	.78510	.44126	.84879	.61522	.28786	.20129	.82686
#1	.40329	10.193	.40176	.48985	1.0165	.05045	10.103
#2	.40723	10.279	.39524	.49189	1.0170	.05064	10.066
#3	.40095	10.211	.40009	.49581	1.0218	.05061	10.226
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05109**	**.20032**	**.50425**	**.51389**	**3.9993**	**49.456**	**1.0271**
Stddev	.00034	.00030	.00243	.00134	.0288	.230	.0087
%RSD	.66912	.15054	.48104	.26170	.72080	.46581	.84751
#1	.05096	.20027	.50150	.51248	3.9697	49.237	1.0193
#2	.05148	.20064	.50519	.51515	4.0273	49.435	1.0365
#3	.05084	.20004	.50607	.51404	4.0011	49.697	1.0255
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.5965**	**.50354**	**1.0171**	**50.205**	**.50852**	**9.9839**	**.50630**
Stddev	.0605	.00125	.0104	.188	.00105	.0190	.00443
%RSD	.63000	.24843	1.0188	.37433	.20655	.18977	.87552
#1	9.5498	.50228	1.0149	50.026	.50872	9.9714	.50177
#2	9.5748	.50478	1.0283	50.189	.50945	10.006	.51063
#3	9.6648	.50354	1.0080	50.400	.50738	9.9747	.50650
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 21:34:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2057**	**.39535**	**5.1622**	**1.0068**	**1.0152**	**1.0163**	**.51868**
Stddev	.0009	.00724	.0066	.0035	.0056	.0047	.00189
%RSD	.07645	1.8317	.12822	.34938	.55282	.46463	.36343
#1	1.2067	.40045	5.1563	1.0056	1.0088	1.0109	.51803
#2	1.2055	.39854	5.1610	1.0107	1.0174	1.0194	.51721
#3	1.2048	.38706	5.1694	1.0040	1.0193	1.0187	.52081
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0297**	**1.0108**	F **2.9654**
Stddev	.0010	.0013	.0124
%RSD	.09618	.12981	.41944
#1	1.0297	1.0109	2.9530
#2	1.0307	1.0120	2.9654
#3	1.0287	1.0094	2.9779
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9765.3**	**59222.**	**3028.4**
Stddev	12.6	103.	19.9
%RSD	.12950	.17379	.65564
#1	9777.2	59331.	3048.9
#2	9752.0	59127.	3027.1
#3	9766.5	59207.	3009.2

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 21:38:07 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00048	-.00056	.00159	.00238	.00173	.00014	-.05704	-.00011
Stddev	.00089	.00372	.00374	.00230	.00036	.00002	.05178	.00017
%RSD	185.89	660.56	234.30	96.559	20.708	14.400	90.783	156.11
#1	-.00133	.00080	.00118	.00024	.00195	.00012	-.11682	-.00027
#2	-.00055	-.00477	-.00192	.00481	.00132	.00016	-.02846	-.00014
#3	.00045	.00228	.00552	.00209	.00192	.00015	-.02584	.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00017	.00048	.00001	.01554	-.08094	.00031	-.00960	.00059
Stddev	.00009	.00037	.00173	.01679	.11260	.00565	.05873	.00171
%RSD	56.610	77.335	20230.	108.07	139.11	1853.2	611.58	288.38
#1	.00024	.00066	.00194	-.00104	.01136	.00158	-.02063	.00007
#2	.00020	.00072	-.00054	.03254	-.20639	-.00588	.05386	.00251
#3	.00006	.00005	-.00138	.01511	-.04780	.00521	-.06203	-.00080
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00336	.02885	-.00050	.00712	.00213	.00602	-.00054	.04214
Stddev	.00024	.02974	.00074	.00312	.00180	.00158	.00069	.00013
%RSD	7.0582	103.10	148.29	43.907	84.214	26.247	127.12	.29697
#1	.00323	-.00548	.00008	.00759	.00359	.00782	.00008	.04201
#2	.00321	.04496	-.00024	.00998	.00013	.00485	-.00042	.04226
#3	.00363	.04705	-.00133	.00378	.00268	.00540	-.00129	.04215
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 21:38:07 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00047**	**.00065**	**-.00177**	**-.00021**	**.00113**	**.00005**	**.00375**
Stddev	.00100	.00044	.00221	.00418	.00029	.00020	.00087
%RSD	212.41	68.047	125.42	1958.4	25.944	435.57	23.135
#1	.00160	.00080	.00075	-.00080	.00139	-.00005	.00437
#2	-.00031	.00015	-.00341	-.00407	.00081	.00028	.00276
#3	.00012	.00100	-.00264	.00422	.00119	-.00009	.00413
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9540.4**	**58573.**	**2934.4**
Stddev	6.2	75.	17.1
%RSD	.06498	.12729	.58178
#1	9533.2	58646.	2919.8
#2	9543.3	58575.	2953.2
#3	9544.6	58497.	2930.3

Approved: June 27, 2016

Sample Name: PBW 9A Acquired: 6/25/2016 21:42:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**.00614**	**.00175**	**.00056**	**.00231**	**.00012**	**-.04646**	**-.00007**
Stddev	.00171	.00204	.00393	.00252	.00071	.00020	.04124	.00026
%RSD	1283.4	33.141	224.26	449.74	30.683	160.24	88.765	365.50
#1	.00182	.00611	.00626	-.00197	.00237	.00018	-.00258	-.00036
#2	-.00137	.00820	-.00088	.00057	.00157	.00029	-.08442	.00014
#3	-.00085	.00413	-.00013	.00307	.00298	-.00010	-.05240	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00037**	**.00010**	**.00080**	**.02407**	**-.12665**	**.00141**	**-.00997**	**-.00231**
Stddev	.00033	.00034	.00081	.01192	.12266	.00378	.02213	.00072
%RSD	90.436	329.64	101.91	49.545	96.850	267.21	222.01	31.042
#1	-.00000	-.00017	-.00014	.03113	-.25310	.00322	.01487	-.00193
#2	-.00045	.00048	.00118	.03078	-.11869	.00395	-.02759	-.00185
#3	-.00065	-.00000	.00135	.01030	-.00816	-.00293	-.01718	-.00313
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00037**	**.03613**	**.00006**	**.00672**	**.00326**	**-.00431**	**-.00184**	**.00566**
Stddev	.00007	.03275	.00077	.00152	.00253	.00485	.00591	.00059
%RSD	18.235	90.627	1212.2	22.673	77.664	112.59	322.28	10.478
#1	.00035	.07347	.00060	.00612	.00546	-.00044	-.00125	.00618
#2	.00032	.02263	-.00082	.00846	.00049	-.00975	.00376	.00501
#3	.00045	.01230	.00041	.00559	.00385	-.00274	-.00802	.00578
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: PBW 9A Acquired: 6/25/2016 21:42:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00063**	**.00033**	**-.00143**	**-.00317**	**.00125**	**.00161**	**.00218**
Stddev	.00053	.00015	.00130	.00329	.00039	.00011	.00081
%RSD	84.270	46.594	90.758	103.97	30.982	7.0758	37.192
#1	.00094	.00050	-.00075	-.00532	.00080	.00148	.00195
#2	.00002	.00020	-.00062	-.00481	.00145	.00171	.00150
#3	.00092	.00028	-.00293	.00062	.00150	.00162	.00308
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9585.8**	**59115.**	**2963.7**
Stddev	26.3	98.	6.0
%RSD	.27487	.16506	.20388
#1	9574.9	59148.	2962.9
#2	9615.8	59005.	2958.2
#3	9566.6	59192.	2970.2

Approved: June 27, 2016

Sample Name: LCSW 9A Acquired: 6/25/2016 21:45:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19278**	**4.8645**	**.19357**	**.92620**	**.49567**	**.02388**	**4.9744**	**.02438**
Stddev	.00059	.0150	.00172	.00296	.00123	.00001	.0131	.00027
%RSD	.30529	.30879	.88618	.31973	.24905	.05168	.26315	1.0959
#1	.19266	4.8773	.19194	.92961	.49521	.02387	4.9745	.02469
#2	.19342	4.8480	.19536	.92439	.49707	.02389	4.9613	.02424
#3	.19226	4.8682	.19339	.92458	.49473	.02389	4.9875	.02421
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09756**	**.24011**	**.24908**	**1.9287**	**23.845**	**.50347**	**4.7458**	**.24452**
Stddev	.00018	.00063	.00057	.0379	.076	.00304	.1100	.00304
%RSD	.18802	.26044	.22811	1.9678	.32025	.60437	2.3184	1.2446
#1	.09771	.24070	.24970	1.9204	23.902	.50653	4.7834	.24200
#2	.09761	.23946	.24858	1.9701	23.875	.50343	4.6218	.24790
#3	.09735	.24018	.24895	1.8956	23.758	.50044	4.8320	.24367
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50361**	**24.550**	**.24652**	**4.6792**	**.24585**	**.57643**	**.18945**	**2.4007**
Stddev	.00079	.045	.00091	.0152	.00374	.00321	.00289	.0038
%RSD	.15600	.18503	.37091	.32489	1.5220	.55613	1.5253	.15739
#1	.50355	24.561	.24748	4.6824	.24789	.57981	.19277	2.4049
#2	.50442	24.588	.24565	4.6926	.24153	.57604	.18748	2.3993
#3	.50285	24.500	.24642	4.6627	.24812	.57343	.18811	2.3978
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LCSW 9A Acquired: 6/25/2016 21:45:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48346**	**.50078**	**.49876**	**.24644**	**.49280**	**.48316**	**.01652**
Stddev	.00157	.00182	.00430	.00290	.00093	.00040	.00416
%RSD	.32431	.36335	.86149	1.1750	.18883	.08190	25.157
#1	.48465	.50112	.49719	.24479	.49339	.48270	.01275
#2	.48404	.50241	.49547	.24978	.49173	.48334	.01583
#3	.48168	.49882	.50362	.24474	.49328	.48342	.02097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9876.1**	**60275.**	**3070.8**
Stddev	6.3	180.	12.0
%RSD	.06355	.29822	.39033
#1	9874.5	60221.	3071.7
#2	9883.0	60476.	3058.4
#3	9870.8	60129.	3082.3

Approved: June 27, 2016

Sample Name: F BLANK Acquired: 6/25/2016 21:49:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00055	-.00126	-.00015	.00486	.00128	.00012	-.00459	.00000
Stddev	.00056	.00100	.00110	.00023	.00095	.00001	.02969	.00020
%RSD	100.53	79.131	736.67	4.6731	74.371	6.7718	646.26	7554.0
#1	.00064	-.00237	-.00130	.00509	.00028	.00012	.02032	.00019
#2	.00106	-.00043	-.00004	.00484	.00217	.00011	-.03745	-.00020
#3	-.00004	-.00099	.00090	.00463	.00139	.00012	.00335	.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00037	.00026	.00088	.03134	-.11502	-.00026	.03535	-.00044
Stddev	.00018	.00121	.00068	.00659	.05233	.00265	.10514	.00096
%RSD	49.613	468.36	76.871	21.019	45.492	1017.6	297.45	218.87
#1	.00058	.00075	.00159	.02374	-.13830	.00130	-.07629	.00060
#2	.00024	.00114	.00082	.03485	-.15168	.00124	.13249	-.00129
#3	.00028	-.00112	.00024	.03543	-.05510	-.00331	.04985	-.00062
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00056	.07370	.00132	.00313	.00117	-.00091	.00077	-.00309
Stddev	.00026	.00588	.00035	.00217	.00118	.00090	.00482	.00173
%RSD	45.817	7.9801	26.636	69.209	100.20	98.397	629.48	56.034
#1	.00085	.06695	.00170	.00515	.00095	-.00029	-.00041	-.00378
#2	.00042	.07645	.00125	.00084	.00245	-.00194	-.00336	-.00438
#3	.00040	.07770	.00101	.00340	.00013	-.00050	.00606	-.00112
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: F BLANK Acquired: 6/25/2016 21:49:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00033	.00302	-.00048	-.00333	.00002	.00245	.00132
Stddev	.00077	.00023	.00928	.00236	.00016	.00009	.00057
%RSD	230.12	7.5634	1914.7	70.982	654.43	3.7966	42.974
#1	.00023	.00325	-.01051	-.00146	-.00016	.00251	.00071
#2	.00115	.00279	.00127	-.00253	.00008	.00234	.00183
#3	-.00038	.00303	.00779	-.00598	.00016	.00250	.00143
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9948.0	61128.	3082.2
Stddev	8.3	35.	19.2
%RSD	.08376	.05747	.62310
#1	9943.7	61136.	3060.8
#2	9942.7	61089.	3097.9
#3	9957.6	61158.	3087.8

Approved: June 27, 2016

Sample Name: L1606078709 Acquired: 6/25/2016 21:53:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00106	.32067	.00199	.03822	.08666	.00012	52.298	.00025
Stddev	.00033	.00225	.00178	.00083	.00065	.00013	.384	.00005
%RSD	31.425	.70249	89.821	2.1841	.74667	104.78	.73398	21.021
#1	.00100	.32316	.00023	.03790	.08700	-.00002	52.414	.00024
#2	.00077	.32010	.00380	.03760	.08707	.00018	51.870	.00021
#3	.00143	.31876	.00193	.03917	.08592	.00021	52.611	.00031
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	.00232	.00233	.31407	.84091	.00625	16.422	.01281
Stddev	.00025	.00123	.00162	.01381	.09733	.00295	.087	.00086
%RSD	177.99	52.851	69.798	4.3981	11.574	47.203	.53048	6.7172
#1	-.00001	.00331	.00164	.32378	.73706	.00531	16.422	.01228
#2	.00000	.00271	.00116	.32018	.93004	.00389	16.510	.01236
#3	.00043	.00095	.00418	.29826	.85562	.00956	16.336	.01381
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00023	3.5331	.00030	.18004	.00409	-.00180	.00164	8.5479
Stddev	.00032	.0286	.00045	.00328	.00158	.00258	.00446	.0078
%RSD	137.02	.80985	147.61	1.8192	38.579	143.43	271.62	.09172
#1	-.00011	3.5554	-.00012	.17874	.00498	-.00359	.00673	8.5558
#2	.00052	3.5008	.00077	.17762	.00503	-.00297	-.00020	8.5477
#3	.00028	3.5432	.00026	.18377	.00227	.00116	-.00161	8.5401
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078709 Acquired: 6/25/2016 21:53:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00032	.13816	-.00281	-.00107	.00371	.00875	.00154
Stddev	.00042	.00078	.00681	.00141	.00081	.00007	.00044
%RSD	132.89	.56285	242.62	131.61	21.926	.80395	28.630
#1	-.00012	.13842	.00500	-.00130	.00416	.00881	.00162
#2	.00035	.13728	-.00586	-.00235	.00277	.00878	.00193
#3	.00072	.13877	-.00757	.00044	.00421	.00867	.00106
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9876.5	60888.	3069.7
Stddev	11.3	192.	3.6
%RSD	.11437	.31559	.11816
#1	9889.0	60702.	3066.7
#2	9866.9	60874.	3073.8
#3	9873.6	61086.	3068.7

Approved: June 27, 2016

Sample Name: L1606078711S Acquired: 6/25/2016 21:57:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19226**	**5.2769**	**.19700**	**.97637**	**.57407**	**.02377**	**55.412**	**.02464**
Stddev	.00232	.0172	.00210	.00695	.00123	.00006	.275	.00018
%RSD	1.2075	.32678	1.0663	.71175	.21357	.23547	.49636	.71677
#1	.19492	5.2928	.19484	.98432	.57271	.02383	55.376	.02446
#2	.19066	5.2586	.19711	.97142	.57439	.02373	55.156	.02465
#3	.19119	5.2792	.19903	.97337	.57510	.02375	55.703	.02481
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09590**	**.23900**	**.24525**	**2.2503**	**24.873**	**.51320**	**20.375**	**.25171**
Stddev	.00047	.00136	.00161	.0156	.173	.01009	.023	.00267
%RSD	.48644	.56990	.65491	.69491	.69517	1.9665	.11329	1.0594
#1	.09633	.24013	.24639	2.2591	25.016	.52369	20.389	.25241
#2	.09595	.23749	.24595	2.2322	24.681	.50356	20.349	.25396
#3	.09541	.23938	.24342	2.2595	24.921	.51234	20.388	.24877
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50427**	**27.841**	**.24087**	**4.9542**	**.24449**	**.58622**	**.18341**	**11.076**
Stddev	.00227	.102	.00131	.0062	.00416	.00262	.00666	.019
%RSD	.45102	.36806	.54487	.12407	1.6996	.44696	3.6299	.17466
#1	.50494	27.874	.24041	4.9608	.24621	.58442	.18291	11.063
#2	.50614	27.726	.24235	4.9486	.23975	.58502	.19031	11.099
#3	.50174	27.923	.23985	4.9531	.24750	.58923	.17702	11.068
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078711S Acquired: 6/25/2016 21:57:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48537**	**.63044**	**.49547**	**.24660**	**.49048**	**.48728**	**.00709**
Stddev	.00041	.00808	.00649	.00259	.00192	.00292	.00024
%RSD	.08494	1.2823	1.3100	1.0500	.39154	.59943	3.4352
#1	.48528	.63362	.48907	.24492	.49247	.49009	.00732
#2	.48501	.63645	.50205	.24958	.48864	.48426	.00711
#3	.48582	.62125	.49531	.24529	.49034	.48750	.00684
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9767.2**	**59939.**	**3081.9**
Stddev	8.4	151.	2.2
%RSD	.08617	.25130	.07156
#1	9776.8	59826.	3083.9
#2	9761.1	60110.	3082.2
#3	9763.8	59880.	3079.5

Approved: June 27, 2016

Sample Name: L1606078712SD Acquired: 6/25/2016 22:01:06 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-06

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19422**	**6.5547**	**.19616**	**.97909**	**.58447**	**.02412**	**57.272**	**.02449**
Stddev	.00043	.0174	.00206	.00320	.00356	.00008	.255	.00025
%RSD	.22056	.26500	1.0502	.32643	.60985	.31860	.44461	1.0216
#1	.19468	6.5722	.19401	.98024	.58211	.02413	57.018	.02471
#2	.19383	6.5546	.19811	.97548	.58857	.02404	57.527	.02422
#3	.19415	6.5375	.19636	.98156	.58273	.02419	57.272	.02454
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09680**	**.24221**	**.24688**	**3.2519**	**25.085**	**.51256**	**21.272**	**.29044**
Stddev	.00024	.00084	.00160	.0305	.123	.00182	.123	.00255
%RSD	.25204	.34673	.64632	.93681	.48984	.35468	.57596	.87927
#1	.09682	.24178	.24538	3.2175	24.965	.51125	21.131	.28769
#2	.09704	.24318	.24856	3.2756	25.081	.51179	21.342	.29274
#3	.09655	.24168	.24672	3.2624	25.210	.51464	21.344	.29089
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50646**	**28.018**	**.24386**	**4.9818**	**.24358**	**.58240**	**.18769**	**13.764**
Stddev	.00076	.167	.00147	.0053	.00228	.00382	.00425	.008
%RSD	.15029	.59670	.60238	.10534	.93711	.65545	2.2629	.06131
#1	.50691	27.825	.24251	4.9763	.24111	.57903	.18298	13.756
#2	.50690	28.112	.24543	4.9823	.24561	.58163	.18886	13.772
#3	.50558	28.117	.24364	4.9868	.24402	.58655	.19123	13.765
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078712SD Acquired: 6/25/2016 22:01:06 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-06

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48752**	**.63979**	**.50344**	**.24632**	**.49636**	**.48329**	**.00682**
Stddev	.00123	.00383	.00416	.00077	.00180	.00044	.00097
%RSD	.25232	.59818	.82653	.31336	.36344	.09176	14.263
#1	.48785	.63646	.50591	.24591	.49538	.48373	.00713
#2	.48616	.64397	.49864	.24584	.49525	.48329	.00761
#3	.48856	.63895	.50578	.24721	.49844	.48284	.00573
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9745.6**	**59867.**	**3081.8**
Stddev	9.9	91.	26.0
%RSD	.10194	.15198	.84468
#1	9734.5	59851.	3110.8
#2	9753.6	59965.	3060.5
#3	9748.6	59785.	3074.0

Approved: June 27, 2016

Sample Name: L1606078701 Acquired: 6/25/2016 22:04:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00206	-.00089	-.00017	.00563	.00151	.00011	.08615	.00015
Stddev	.00234	.00167	.00131	.00160	.00060	.00001	.02054	.00021
%RSD	113.42	187.14	751.31	28.470	39.897	9.3705	23.846	140.35
#1	.00316	-.00269	.00118	.00529	.00196	.00012	.07852	-.00008
#2	-.00062	-.00061	-.00144	.00737	.00083	.00010	.07052	.00033
#3	.00364	.00062	-.00027	.00422	.00173	.00012	.10942	.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00013	.00093	.00322	.06115	.01246	.00241	.02800	.00003
Stddev	.00003	.00076	.00095	.00519	.08806	.00216	.01589	.00083
%RSD	20.822	82.163	29.586	8.4819	706.92	89.891	56.731	2625.4
#1	.00011	.00029	.00408	.06573	-.03307	.00026	.04634	-.00085
#2	.00017	.00177	.00219	.05552	.11396	.00458	.01831	.00015
#3	.00013	.00071	.00338	.06220	-.04352	.00238	.01936	.00079
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00040	.44862	.00009	.00157	.00319	-.00012	.00644	.10535
Stddev	.00015	.01169	.00041	.00200	.00252	.00436	.00247	.00160
%RSD	36.378	2.6064	429.23	127.80	78.785	3632.8	38.299	1.5199
#1	.00041	.45408	.00056	-.00069	.00238	-.00321	.00918	.10450
#2	.00025	.45659	-.00021	.00226	.00118	.00487	.00438	.10720
#3	.00054	.43520	-.00006	.00313	.00602	-.00202	.00577	.10436
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078701 Acquired: 6/25/2016 22:04:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00114**	**.00134**	**.00464**	**-.00236**	**.00126**	**.00852**	**.00204**
Stddev	.00023	.00031	.01117	.00170	.00022	.00013	.00133
%RSD	19.939	23.344	240.52	72.025	17.171	1.4749	64.993
#1	.00100	.00168	-.00728	-.00407	.00147	.00841	.00183
#2	.00140	.00106	.01487	-.00232	.00127	.00866	.00083
#3	.00102	.00128	.00634	-.00068	.00104	.00851	.00346
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9978.6**	**61725.**	**3108.7**
Stddev	25.8	52.	12.8
%RSD	.25894	.08436	.41157
#1	9949.5	61751.	3109.2
#2	9998.7	61759.	3121.3
#3	9987.8	61665.	3095.7

Approved: June 27, 2016

Sample Name: L1606078701PS Acquired: 6/25/2016 22:08:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573561-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19562**	**4.8670**	**.19417**	**.93274**	**.49617**	**.02396**	**5.0466**	**.02449**
Stddev	.00101	.0014	.00206	.00077	.00126	.00015	.0443	.00024
%RSD	.51610	.02767	1.0615	.08264	.25431	.60837	.87718	.96002
#1	.19561	4.8655	.19531	.93216	.49577	.02379	5.0119	.02471
#2	.19461	4.8672	.19179	.93245	.49515	.02407	5.0964	.02451
#3	.19663	4.8682	.19541	.93362	.49758	.02401	5.0313	.02424
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09815**	**.24178**	**.25186**	**1.9475**	**23.902**	**.50912**	**4.6690**	**.24615**
Stddev	.00038	.00032	.00085	.0476	.079	.00280	.1745	.00176
%RSD	.38848	.13325	.33570	2.4417	.33116	.54926	3.7379	.71493
#1	.09847	.24215	.25095	2.0020	23.984	.51180	4.5199	.24606
#2	.09825	.24158	.25263	1.9144	23.825	.50622	4.6262	.24443
#3	.09773	.24160	.25200	1.9261	23.897	.50934	4.8609	.24795
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50547**	**24.979**	**.24871**	**4.7191**	**.25003**	**.58513**	**.18331**	**2.5397**
Stddev	.00148	.047	.00145	.0045	.00153	.00133	.00963	.0085
%RSD	.29370	.18695	.58415	.09482	.61173	.22744	5.2541	.33396
#1	.50679	24.962	.24995	4.7237	.24937	.58423	.17385	2.5470
#2	.50576	24.943	.24906	4.7186	.24895	.58665	.18298	2.5417
#3	.50387	25.032	.24711	4.7148	.25178	.58449	.19310	2.5304
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078701PS Acquired: 6/25/2016 22:08:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573561-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48806**	**.50437**	**.49939**	**.24983**	**.48939**	**.48880**	**.00250**
Stddev	.00273	.00349	.00591	.00248	.00296	.00032	.00045
%RSD	.55970	.69122	1.1836	.99157	.60527	.06447	18.096
#1	.48918	.50154	.50066	.25202	.49256	.48872	.00242
#2	.49005	.50331	.49295	.25032	.48893	.48914	.00298
#3	.48494	.50827	.50457	.24714	.48669	.48853	.00209
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9876.1**	**60662.**	**3087.6**
Stddev	12.0	38.	4.3
%RSD	.12163	.06253	.13975
#1	9869.1	60619.	3083.6
#2	9869.3	60677.	3092.2
#3	9890.0	60690.	3087.1

Approved: June 27, 2016

Sample Name: L1606078701SDL Acquired: 6/25/2016 22:12:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573561-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00226**	**-.00962**	**.00257**	**.00372**	**.00008**	**.00010**	**-.05467**	**-.00005**
Stddev	.00390	.01062	.00091	.00131	.00131	.00006	.03854	.00004
%RSD	172.35	110.37	35.549	35.155	1747.1	57.750	70.491	84.699
#1	.00675	-.01272	.00350	.00225	-.00089	.00014	-.04307	-.00000
#2	.00027	-.01835	.00254	.00475	.00157	.00012	-.09767	-.00007
#3	-.00023	.00220	.00167	.00417	-.00046	.00003	-.02326	-.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00015**	**.00076**	**.00146**	**.02792**	**-.10756**	**-.00176**	**-.01577**	**.00085**
Stddev	.00022	.00111	.00071	.01781	.06853	.00352	.01010	.00145
%RSD	144.06	146.16	48.411	63.806	63.714	199.76	64.076	169.31
#1	.00010	.00182	.00216	.04804	-.06727	-.00153	-.01532	-.00005
#2	-.00026	-.00038	.00146	.01416	-.18668	.00163	-.02609	.00009
#3	-.00029	.00083	.00075	.02156	-.06871	-.00539	-.00589	.00252
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00035**	**.12025**	**-.00037**	**.00742**	**.00388**	**-.00036**	**.00693**	**.02501**
Stddev	.00028	.02622	.00009	.00699	.00334	.00491	.00079	.00132
%RSD	80.313	21.803	23.237	94.157	86.143	1346.9	11.341	5.2920
#1	.00057	.14797	-.00041	.00450	.00651	.00107	.00746	.02626
#2	.00045	.09585	-.00027	.01539	.00501	-.00584	.00603	.02362
#3	.00003	.11693	-.00042	.00237	.00012	.00367	.00731	.02515
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078701SDL Acquired: 6/25/2016 22:12:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573561-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00036	.00027	.00391	-.00239	.00091	.00198	.00117
Stddev	.00062	.00058	.00259	.00083	.00057	.00021	.00105
%RSD	170.80	217.03	66.210	34.804	62.977	10.839	89.133
#1	.00032	-.00015	.00203	-.00335	.00041	.00217	.00108
#2	-.00024	.00003	.00686	-.00188	.00153	.00175	.00018
#3	.00100	.00093	.00284	-.00194	.00079	.00202	.00227
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10481.	64804.	3221.0
Stddev	25.	98.	4.2
%RSD	.23393	.15113	.13029
#1	10501.	64916.	3225.8
#2	10489.	64737.	3219.2
#3	10454.	64758.	3217.9

Approved: June 27, 2016

Sample Name: L1606078701SDL Acquired: 6/25/2016 22:16:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573561-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00050**	**-.01671**	**.00066**	**.00174**	**.00117**	**.00007**	**-.03873**	**-.00006**
Stddev	.00055	.00595	.00077	.00373	.00047	.00003	.02526	.00023
%RSD	110.50	35.627	115.52	214.66	39.929	36.665	65.224	418.10
#1	.00087	-.01063	-.00016	.00037	.00065	.00010	-.06018	.00018
#2	-.00014	-.02253	.00136	-.00111	.00156	.00005	-.01089	-.00029
#3	.00077	-.01699	.00080	.00596	.00129	.00006	-.04513	-.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00020**	**-.00027**	**.00015**	**.01537**	**-.08473**	**-.00313**	**-.04452**	**.00037**
Stddev	.00041	.00031	.00072	.00667	.12317	.00148	.02793	.00217
%RSD	208.97	115.52	489.41	43.427	145.37	47.326	62.733	580.66
#1	.00051	-.00025	-.00010	.01465	-.11440	-.00155	-.01300	-.00212
#2	-.00027	-.00059	-.00042	.02238	.05057	-.00335	-.06618	.00176
#3	.00035	.00003	.00096	.00908	-.19035	-.00448	-.05440	.00148
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00052**	**.04068**	**-.00002**	**.00630**	**.00223**	**-.00439**	**.00789**	**.00775**
Stddev	.00014	.01560	.00098	.00605	.00105	.00119	.00333	.00034
%RSD	26.596	38.347	4828.5	96.010	47.199	27.128	42.126	4.3419
#1	-.00068	.05533	-.00110	-.00065	.00121	-.00319	.01093	.00736
#2	-.00042	.02428	.00080	.00920	.00218	-.00557	.00434	.00794
#3	-.00046	.04243	.00023	.01036	.00331	-.00443	.00841	.00794
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078701SDL Acquired: 6/25/2016 22:16:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573561-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00078**	**.00049**	**.00611**	**-.00169**	**-.00019**	**.00076**	**.00104**
Stddev	.00047	.00028	.00364	.00121	.00086	.00008	.00084
%RSD	59.598	55.989	59.553	71.357	443.84	10.636	80.846
#1	-.00025	.00020	.00285	-.00078	-.00105	.00067	.00138
#2	-.00100	.00075	.00545	-.00306	-.00021	.00078	.00165
#3	-.00110	.00052	.01004	-.00123	.00067	.00082	.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10473.**	**64343.**	**3185.8**
Stddev	11.	88.	33.4
%RSD	.10369	.13736	1.0486
#1	10479.	64432.	3210.0
#2	10479.	64255.	3199.9
#3	10461.	64342.	3147.7

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 22:20:12 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.39676**	**10.011**	**.39200**	**.48774**	**1.0072**	**.04931**	**9.9627**
Stddev	.00166	.042	.00060	.00247	.0020	.00016	.0336
%RSD	.41961	.41526	.15324	.50705	.19487	.32345	.33737
#1	.39852	10.036	.39268	.48499	1.0052	.04934	9.9285
#2	.39656	10.035	.39180	.48843	1.0074	.04945	9.9639
#3	.39521	9.9633	.39153	.48979	1.0091	.04914	9.9956
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05046**	**.19748**	**.49159**	**.50549**	**3.9250**	**48.860**	**1.0117**
Stddev	.00033	.00063	.00167	.00263	.0491	.051	.0023
%RSD	.66275	.32062	.34000	.52017	1.2522	.10387	.22425
#1	.05068	.19812	.49231	.50540	3.8948	48.907	1.0140
#2	.05064	.19745	.49279	.50817	3.9817	48.806	1.0118
#3	.05008	.19686	.48968	.50291	3.8986	48.867	1.0095
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.4449**	**.49486**	**1.0080**	**49.891**	**.50156**	**9.8436**	**.50078**
Stddev	.0163	.00095	.0079	.070	.00172	.0350	.00128
%RSD	.17297	.19183	.78730	.13958	.34250	.35531	.25500
#1	9.4275	.49541	1.0170	49.848	.50286	9.8791	.50070
#2	9.4600	.49376	1.0019	49.971	.50221	9.8426	.49954
#3	9.4471	.49540	1.0052	49.853	.49961	9.8092	.50209
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 22:20:12 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1892**	**.38968**	**5.0794**	**.99180**	**1.0052**	**1.0145**	**.50829**
Stddev	.0070	.00637	.0139	.00335	.0009	.0065	.00390
%RSD	.58775	1.6349	.27365	.33789	.09319	.64413	.76634
#1	1.1970	.39094	5.0928	.99452	1.0044	1.0072	.50887
#2	1.1869	.38277	5.0804	.99283	1.0062	1.0197	.51187
#3	1.1836	.39532	5.0651	.98805	1.0048	1.0166	.50414
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0022**	**.99700**	F **2.9158**
Stddev	.0042	.00301	.0038
%RSD	.42153	.30191	.13010
#1	1.0066	.99806	2.9128
#2	1.0017	.99934	2.9201
#3	.99824	.99361	2.9145
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9930.6**	**60612.**	**3086.3**
Stddev	10.1	73.	9.8
%RSD	.10207	.12111	.31687
#1	9941.9	60647.	3097.4
#2	9922.1	60528.	3082.4
#3	9927.9	60661.	3079.1

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 22:23:49 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00100**	**-.00695**	**.00023**	**.00246**	**.00203**	**.00014**	**-.01220**	**.00014**
Stddev	.00181	.00287	.00117	.00195	.00133	.00002	.03027	.00029
%RSD	181.13	41.303	500.29	79.096	65.555	16.180	248.16	212.80
#1	-.00109	-.00504	-.00098	.00360	.00305	.00017	-.03745	.00025
#2	.00205	-.00555	.00136	.00021	.00053	.00014	.02136	-.00019
#3	.00203	-.01025	.00032	.00357	.00251	.00013	-.02050	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00014**	**-.00000**	**.00097**	**.02954**	**-.02965**	**-.00129**	**.00088**	**.00163**
Stddev	.00011	.00082	.00136	.00883	.01713	.00234	.03199	.00271
%RSD	74.595	562770.	140.78	29.900	57.784	181.57	3634.1	165.69
#1	.00008	.00083	.00213	.03726	-.04191	-.00397	.03754	.00233
#2	.00008	-.00001	.00130	.01991	-.03697	-.00018	-.01355	.00393
#3	.00027	-.00081	-.00053	.03145	-.01007	.00029	-.02135	-.00135
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00354**	**.01474**	**-.00023**	**.00359**	**.00237**	**.00219**	**.00418**	**.03627**
Stddev	.00058	.04847	.00031	.00245	.00060	.00526	.00484	.00058
%RSD	16.314	328.86	136.11	68.357	25.126	240.59	115.86	1.6055
#1	.00288	-.00030	-.00040	.00322	.00284	.00365	.00822	.03616
#2	.00384	.06894	.00013	.00134	.00170	.00656	.00551	.03575
#3	.00391	-.02443	-.00042	.00620	.00257	-.00365	-.00119	.03690
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 22:23:49 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00016**	**.00039**	**.00181**	**-.00370**	**.00020**	**.00004**	**.00435**
Stddev	.00110	.00025	.00867	.00085	.00051	.00012	.00191
%RSD	689.85	64.301	479.78	22.978	253.56	271.57	43.946
#1	-.00106	.00012	.00915	-.00439	-.00031	.00014	.00375
#2	.00108	.00042	.00404	-.00397	.00020	.00008	.00281
#3	.00046	.00061	-.00776	-.00275	.00071	-.00009	.00649
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9799.0**	**60424.**	**3020.0**
Stddev	11.5	135.	11.0
%RSD	.11760	.22419	.36496
#1	9789.9	60445.	3025.5
#2	9795.1	60547.	3027.2
#3	9812.0	60279.	3007.3

Approved: June 27, 2016

Sample Name: L1606078710 Acquired: 6/25/2016 22:27:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00017	.22720	.00214	.03806	.08552	.00012	52.208	.00022
Stddev	.00078	.00436	.00063	.00014	.00079	.00005	.232	.00016
%RSD	450.39	1.9170	29.562	.37345	.92041	43.173	.44436	71.173
#1	.00025	.22486	.00206	.03796	.08534	.00008	52.380	.00034
#2	.00030	.22452	.00280	.03822	.08638	.00018	52.301	.00028
#3	-.00108	.23223	.00155	.03799	.08484	.00010	51.944	.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00002	.00186	.00266	.23670	.85794	.00848	16.196	.00759
Stddev	.00014	.00018	.00107	.03451	.05555	.00209	.053	.00200
%RSD	849.17	9.9559	40.055	14.580	6.4743	24.705	.32764	26.366
#1	-.00011	.00187	.00273	.22408	.80379	.01054	16.251	.00990
#2	.00017	.00167	.00156	.27574	.91479	.00636	16.192	.00643
#3	-.00001	.00204	.00369	.21027	.85526	.00853	16.145	.00644
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00109	3.4758	.00069	.17816	.00790	-.00197	-.00080	8.3609
Stddev	.00043	.0348	.00145	.00114	.00265	.00100	.00335	.0283
%RSD	39.553	1.0002	208.84	.63768	33.519	50.807	416.39	.33868
#1	.00115	3.5101	.00211	.17685	.00777	-.00150	.00187	8.3874
#2	.00149	3.4766	.00076	.17877	.00531	-.00312	.00028	8.3643
#3	.00063	3.4406	-.00079	.17886	.01060	-.00129	-.00456	8.3311
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078710 Acquired: 6/25/2016 22:27:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00106	.13692	-.00198	-.00343	.00351	.00457	.00538
Stddev	.00037	.00076	.00399	.00152	.00135	.00005	.00158
%RSD	34.945	.55142	201.79	44.378	38.390	1.0573	29.298
#1	.00117	.13773	-.00276	-.00427	.00288	.00452	.00538
#2	.00136	.13624	.00235	-.00167	.00505	.00462	.00381
#3	.00065	.13680	-.00552	-.00433	.00259	.00458	.00696
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9837.0	60854.	3075.5
Stddev	21.0	87.	13.2
%RSD	.21328	.14286	.42972
#1	9813.4	60881.	3070.1
#2	9853.7	60757.	3065.8
#3	9843.8	60924.	3090.5

Approved: June 27, 2016

Sample Name: L1606078718 Acquired: 6/25/2016 22:31:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00018	1.2702	.00086	.07193	.15344	.00016	53.535	.00010
Stddev	.00088	.0133	.00129	.00289	.00068	.00006	.315	.00014
%RSD	504.87	1.0437	151.37	4.0242	.44552	39.145	.58814	143.56
#1	.00078	1.2717	-.00017	.07191	.15299	.00023	53.705	-.00002
#2	-.00096	1.2563	.00043	.07484	.15310	.00014	53.172	.00025
#3	-.00034	1.2827	.00231	.06905	.15423	.00011	53.729	.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00093	.00320	.00404	1.3343	1.2661	.00990	20.655	.10718
Stddev	.00038	.00017	.00073	.0192	.0597	.00249	.173	.00269
%RSD	40.455	5.3029	17.991	1.4379	4.7192	25.172	.83716	2.5057
#1	.00050	.00331	.00423	1.3126	1.1986	.01278	20.828	.10559
#2	.00122	.00328	.00465	1.3416	1.2875	.00842	20.482	.11029
#3	.00107	.00300	.00324	1.3489	1.3123	.00851	20.655	.10568
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00197	9.5571	.00271	.15727	.00713	-.00001	-.00213	9.2004
Stddev	.00025	.0745	.00060	.00312	.00099	.00396	.00180	.0136
%RSD	12.754	.77965	22.028	1.9854	13.840	44119.	84.428	.14793
#1	.00222	9.6099	.00232	.16018	.00745	-.00238	-.00366	9.2098
#2	.00198	9.4719	.00242	.15397	.00602	-.00222	-.00015	9.2066
#3	.00172	9.5895	.00340	.15765	.00792	.00457	-.00260	9.1848
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078718 Acquired: 6/25/2016 22:31:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00007**	**.21832**	**.00935**	**-.00138**	**.00550**	**.00807**	**.00587**
Stddev	.00077	.00070	.00368	.00189	.00037	.00008	.00139
%RSD	1050.3	.31961	39.319	137.43	6.8010	.96236	23.773
#1	.00067	.21839	.00655	-.00216	.00508	.00811	.00494
#2	-.00003	.21758	.00798	.00078	.00561	.00812	.00519
#3	-.00086	.21897	.01351	-.00275	.00580	.00798	.00747
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9846.3**	**60728.**	**3091.9**
Stddev	3.6	208.	32.5
%RSD	.03671	.34247	1.0518
#1	9850.4	60702.	3069.6
#2	9844.9	60535.	3129.2
#3	9843.6	60948.	3076.8

Approved: June 27, 2016

Sample Name: L1606078719 Acquired: 6/25/2016 22:35:34 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00022	-.00125	.00081	-.00052	.00163	.00015	.05700	.00032
Stddev	.00275	.01314	.00222	.00254	.00035	.00003	.05177	.00012
%RSD	1222.4	1048.7	273.08	489.15	21.203	21.890	90.827	38.965
#1	-.00200	-.00415	-.00134	.00209	.00158	.00017	.05365	.00040
#2	.00294	.01310	.00068	-.00066	.00200	.00012	.00698	.00018
#3	-.00162	-.01271	.00310	-.00298	.00131	.00018	.11035	.00038
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00006	.00052	.00450	.02391	-.05991	-.00020	.04697	-.00036
Stddev	.00016	.00162	.00035	.01615	.07304	.00309	.12538	.00213
%RSD	276.80	313.97	7.6965	67.537	121.91	1526.0	266.94	591.72
#1	.00015	.00207	.00475	.04225	-.11673	.00336	-.03849	-.00282
#2	-.00012	.00065	.00464	.01187	-.08546	-.00225	-.01151	.00086
#3	.00015	-.00117	.00410	.01759	.02247	-.00171	.19090	.00088
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00030	.25593	-.00012	.00090	.00315	-.00362	.00327	.05573
Stddev	.00023	.02584	.00048	.00460	.00186	.00138	.00112	.00365
%RSD	74.483	10.098	393.31	510.33	59.051	38.233	34.210	6.5471
#1	.00052	.27908	.00017	.00590	.00100	-.00516	.00452	.05848
#2	.00007	.22805	.00014	-.00005	.00421	-.00248	.00234	.05711
#3	.00032	.26066	-.00068	-.00315	.00425	-.00323	.00296	.05159
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078719 Acquired: 6/25/2016 22:35:34 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00098**	**.00207**	**-.00026**	**-.00547**	**.00052**	**.00525**	**.00241**
Stddev	.00027	.00008	.00293	.00052	.00032	.00008	.00097
%RSD	27.115	4.0421	1136.5	9.5358	61.869	1.5786	40.264
#1	.00096	.00217	-.00025	-.00546	.00057	.00519	.00232
#2	.00073	.00203	-.00319	-.00495	.00081	.00521	.00342
#3	.00126	.00203	.00267	-.00599	.00017	.00534	.00149
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9914.0**	**61492.**	**3065.3**
Stddev	9.7	119.	10.3
%RSD	.09735	.19352	.33672
#1	9905.1	61523.	3061.5
#2	9924.3	61360.	3057.3
#3	9912.7	61592.	3076.9

Approved: June 27, 2016

Sample Name: L1606078709 Acquired: 6/25/2016 22:39:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00063	.01725	.00020	.00148	.00412	.00007	2.5193	.00009
Stddev	.00160	.00300	.00146	.00122	.00055	.00007	.0273	.00025
%RSD	256.33	17.406	733.70	82.744	13.249	105.70	1.0828	269.34
#1	-.00115	.01430	.00189	.00205	.00368	.00003	2.5430	-.00015
#2	.00106	.02030	-.00065	.00007	.00396	.00002	2.4895	.00008
#3	.00197	.01716	-.00064	.00231	.00474	.00015	2.5253	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00035	.00087	.00073	.04349	.03288	.00266	.76988	-.00075
Stddev	.00037	.00063	.00114	.01165	.16835	.00415	.07900	.00106
%RSD	107.35	72.316	156.53	26.779	512.03	155.84	10.261	141.15
#1	.00070	.00029	.00142	.03927	-.06801	.00606	.83027	.00043
#2	-.00004	.00153	.00135	.03454	-.06058	.00390	.79891	-.00163
#3	.00039	.00078	-.00059	.05665	.22723	-.00196	.68048	-.00106
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00016	.20336	.00039	.01881	.00195	-.00218	-.00099	.41497
Stddev	.00052	.02965	.00044	.00286	.00086	.00256	.00440	.00248
%RSD	323.08	14.579	115.44	15.203	44.305	117.44	445.87	.59653
#1	-.00058	.19785	-.00012	.01994	.00207	-.00390	.00409	.41320
#2	.00043	.23538	.00054	.02093	.00103	-.00339	-.00379	.41391
#3	-.00033	.17686	.00073	.01556	.00275	.00076	-.00326	.41779
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078709 Acquired: 6/25/2016 22:39:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**.00684**	**.01005**	**.00002**	**.00088**	**.00094**	**.00070**
Stddev	.00032	.00012	.00959	.00200	.00079	.00012	.00147
%RSD	375.67	1.7762	95.434	11093.	88.946	12.687	210.40
#1	.00028	.00687	.01212	-.00222	.00129	.00101	.00212
#2	.00025	.00693	-.00041	.00064	.00138	.00101	.00079
#3	-.00028	.00670	.01844	.00163	-.00002	.00080	-.00081
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10438.**	**64428.**	**3210.2**
Stddev	16.	116.	22.5
%RSD	.15594	.18047	.70191
#1	10456.	64522.	3184.5
#2	10433.	64464.	3226.3
#3	10424.	64298.	3219.8

Approved: June 27, 2016

Sample Name: L1606078711S Acquired: 6/25/2016 22:43:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00901	.25091	.00908	.04716	.02857	.00129	2.7206	.00112
Stddev	.00226	.00893	.00272	.00090	.00098	.00003	.0190	.00018
%RSD	25.135	3.5576	29.953	1.9149	3.4212	2.0076	.69811	15.952
#1	.00654	.24156	.01090	.04731	.02948	.00128	2.7394	.00096
#2	.00949	.25183	.00595	.04798	.02754	.00132	2.7210	.00131
#3	.01099	.25934	.01039	.04620	.02870	.00128	2.7014	.00109
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00484	.01138	.01241	.12881	1.1056	.02562	.95895	.01155
Stddev	.00047	.00095	.00054	.02178	.0733	.00325	.08330	.00186
%RSD	9.6843	8.3365	4.3248	16.912	6.6280	12.671	8.6861	16.074
#1	.00430	.01246	.01288	.12748	1.1086	.02193	.86824	.01328
#2	.00510	.01069	.01253	.10772	1.1774	.02802	1.0320	.00959
#3	.00513	.01099	.01183	.15123	1.0309	.02692	.97661	.01179
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02077	1.4247	.01204	.24038	.01720	.02493	.01091	.53474
Stddev	.00067	.0371	.00126	.00484	.00256	.00194	.00301	.00121
%RSD	3.2317	2.6069	10.481	2.0127	14.893	7.7710	27.534	.22565
#1	.02144	1.3934	.01127	.24412	.01465	.02673	.00768	.53472
#2	.02075	1.4149	.01350	.23492	.01719	.02518	.01362	.53595
#3	.02010	1.4657	.01136	.24211	.01977	.02288	.01143	.53354
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078711S Acquired: 6/25/2016 22:43:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02007**	**.03079**	**.02489**	**.01057**	**.02372**	**.02477**	**.00180**
Stddev	.00062	.00022	.00917	.00037	.00088	.00015	.00112
%RSD	3.0692	.70951	36.838	3.4629	3.6971	.59147	62.402
#1	.02076	.03087	.02809	.01029	.02425	.02492	.00062
#2	.01959	.03097	.01455	.01045	.02271	.02462	.00193
#3	.01984	.03055	.03204	.01098	.02420	.02478	.00285
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10426.**	**64070.**	**3185.9**
Stddev	11.	72.	10.0
%RSD	.10672	.11272	.31382
#1	10417.	64076.	3184.7
#2	10423.	64138.	3176.5
#3	10439.	63994.	3196.4

Approved: June 27, 2016

Sample Name: L1606078712SD Acquired: 6/25/2016 22:47:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-06

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00985	.31033	.01141	.04692	.02793	.00122	2.7514	.00123
Stddev	.00217	.01120	.00035	.00183	.00030	.00009	.0322	.00017
%RSD	21.999	3.6078	3.0998	3.8902	1.0783	7.5455	1.1711	13.538
#1	.00744	.30133	.01139	.04489	.02771	.00131	2.7147	.00104
#2	.01165	.32286	.01106	.04746	.02780	.00124	2.7644	.00130
#3	.01046	.30678	.01177	.04842	.02827	.00112	2.7751	.00135
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00442	.01187	.01326	.16770	1.1233	.02635	1.0271	.01377
Stddev	.00036	.00059	.00082	.01226	.1089	.00034	.0701	.00235
%RSD	8.1823	4.9929	6.1530	7.3088	9.6965	1.2889	6.8269	17.054
#1	.00413	.01121	.01277	.16233	1.0153	.02615	1.0429	.01385
#2	.00482	.01235	.01421	.15905	1.2332	.02617	1.0881	.01607
#3	.00430	.01207	.01281	.18173	1.1215	.02675	.95049	.01138
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02146	1.3700	.01235	.24640	.01637	.02615	.01276	.67342
Stddev	.00028	.0367	.00082	.00464	.00030	.00374	.00186	.00470
%RSD	1.2904	2.6782	6.6457	1.8816	1.8039	14.304	14.555	.69838
#1	.02172	1.3476	.01220	.24696	.01603	.02416	.01223	.67004
#2	.02148	1.3500	.01162	.25073	.01656	.02383	.01483	.67143
#3	.02117	1.4123	.01324	.24151	.01651	.03046	.01123	.67879
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078712SD Acquired: 6/25/2016 22:47:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-06

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02141**	**.03084**	**.02074**	**.01089**	**.02357**	**.02510**	**.00112**
Stddev	.00030	.00031	.00329	.00303	.00061	.00018	.00144
%RSD	1.3990	1.0105	15.854	27.843	2.5684	.72342	128.59
#1	.02165	.03055	.02352	.00739	.02416	.02497	-.00034
#2	.02107	.03117	.01711	.01260	.02359	.02501	.00117
#3	.02151	.03078	.02157	.01268	.02295	.02530	.00254
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10421.**	**63887.**	**3174.5**
Stddev	13.	153.	11.8
%RSD	.12017	.23969	.37110
#1	10421.	64063.	3186.8
#2	10434.	63786.	3163.4
#3	10409.	63811.	3173.3

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 22:51:13 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.39829**	**10.014**	**.39328**	**.48763**	**1.0098**	**.04913**	**9.9981**
Stddev	.00170	.014	.00352	.00156	.0114	.00024	.0222
%RSD	.42599	.13477	.89524	.31896	1.1283	.49183	.22224
#1	.39634	10.025	.39615	.48872	1.0228	.04926	10.024
#2	.39905	9.9992	.38935	.48585	1.0049	.04885	9.9823
#3	.39946	10.019	.39433	.48833	1.0016	.04927	9.9886
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05040**	**.19854**	**.49076**	**.50671**	**3.9457**	**49.325**	**1.0128**
Stddev	.00023	.00052	.00245	.00059	.0608	.368	.0109
%RSD	.45580	.25959	.49941	.11597	1.5404	.74655	1.0755
#1	.05026	.19800	.48916	.50736	4.0058	49.743	1.0249
#2	.05027	.19858	.48954	.50622	3.9470	49.185	1.0101
#3	.05066	.19903	.49358	.50654	3.8842	49.047	1.0036
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.4894**	**.49712**	**1.0151**	**50.056**	**.50217**	**9.8694**	**.50095**
Stddev	.2168	.00232	.0035	.479	.00145	.0113	.00387
%RSD	2.2847	.46695	.34803	.95643	.28811	.11418	.77216
#1	9.7395	.49968	1.0186	50.588	.50383	9.8565	.50471
#2	9.3739	.49655	1.0152	49.919	.50140	9.8745	.49698
#3	9.3548	.49515	1.0115	49.660	.50126	9.8772	.50116
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 22:51:13 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1957**	**.39554**	**5.0959**	**.99417**	**1.0077**	**1.0177**	**.50994**
Stddev	.0055	.00372	.0051	.00169	.0107	.0101	.00264
%RSD	.46066	.94108	.09932	.17029	1.0641	.98980	.51770
#1	1.1919	.39655	5.0908	.99223	1.0196	1.0259	.51299
#2	1.2020	.39866	5.0958	.99500	1.0048	1.0064	.50861
#3	1.1932	.39142	5.1010	.99530	.99871	1.0207	.50824
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.99906**	**1.0000**	F **2.9202**
Stddev	.00128	.0004	.0305
%RSD	.12818	.03891	1.0450
#1	1.0005	1.0005	2.9552
#2	.99795	.99972	2.9065
#3	.99877	.99993	2.8990
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9918.3**	**60256.**	**3072.6**
Stddev	6.2	156.	34.9
%RSD	.06223	.25822	1.1374
#1	9919.7	60085.	3032.6
#2	9923.7	60388.	3088.2
#3	9911.6	60295.	3097.1

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 22:54:51 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00016**	**.00017**	**.00165**	**.00199**	**.00201**	**.00011**	**-.05290**	**-.00006**
Stddev	.00211	.00653	.00321	.00079	.00009	.00008	.03288	.00016
%RSD	1282.1	3936.0	193.87	39.670	4.4194	70.107	62.164	240.48
#1	-.00117	.00347	.00046	.00135	.00191	.00015	-.05612	-.00004
#2	.00226	-.00735	.00528	.00288	.00202	.00015	-.08405	.00008
#3	-.00158	.00438	-.00078	.00176	.00209	.00002	-.01852	-.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00001**	**.00100**	**.00120**	**.03069**	**-.03948**	**.00003**	**-.00536**	**.00015**
Stddev	.00013	.00044	.00061	.00550	.07927	.00292	.09167	.00308
%RSD	995.60	43.627	50.884	17.919	200.80	8764.3	1709.0	2010.9
#1	-.00014	.00145	.00060	.03538	.05200	-.00061	-.10767	.00091
#2	.00000	.00096	.00116	.03205	-.08238	-.00251	.06932	.00278
#3	.00011	.00058	.00182	.02464	-.08805	.00322	.02226	-.00324
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00351**	**.02455**	**-.00057**	**.00388**	**.00014**	**.00562**	**.00087**	**.03540**
Stddev	.00076	.01905	.00055	.00530	.00438	.00343	.00632	.00172
%RSD	21.717	77.584	95.045	136.67	3025.0	61.032	726.47	4.8534
#1	.00266	.03539	-.00120	-.00182	.00487	.00872	.00338	.03589
#2	.00374	.03571	-.00017	.00867	-.00067	.00193	.00554	.03683
#3	.00413	.00256	-.00036	.00480	-.00377	.00622	-.00632	.03349
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 22:54:51 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00099**	**.00045**	**.00606**	**-.00120**	**.00030**	**-.00008**	**.00363**
Stddev	.00065	.00027	.00156	.00179	.00040	.00006	.00038
%RSD	65.956	60.056	25.746	148.69	132.31	80.432	10.554
#1	.00115	.00019	.00786	-.00316	.00076	-.00006	.00364
#2	.00155	.00073	.00529	-.00079	.00014	-.00015	.00325
#3	.00027	.00044	.00504	.00034	.00001	-.00003	.00401
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9678.8**	**59649.**	**2985.7**
Stddev	6.9	105.	9.5
%RSD	.07084	.17603	.31957
#1	9671.2	59575.	2994.0
#2	9681.0	59769.	2975.3
#3	9684.3	59604.	2987.7

Approved: June 27, 2016

Sample Name: PBW LJ Acquired: 6/25/2016 22:58:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00058**	**-.00592**	**.00193**	**-.00086**	**.00070**	**.00018**	**-.07344**	**-.00023**
Stddev	.00178	.00410	.00330	.00167	.00093	.00007	.04096	.00010
%RSD	307.62	69.294	170.45	194.42	133.02	36.643	55.774	43.143
#1	-.00092	-.00625	.00300	-.00003	-.00002	.00015	-.02635	-.00025
#2	-.00217	-.00166	.00456	.00023	.00174	.00026	-.10082	-.00033
#3	.00135	-.00985	-.00176	-.00279	.00036	.00013	-.09313	-.00013
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00016**	**.00069**	**.00074**	**.03692**	**-.07694**	**.00146**	**.06376**	**.00092**
Stddev	.00002	.00038	.00020	.01230	.06836	.00025	.04368	.00067
%RSD	10.723	54.599	26.935	33.312	88.851	16.877	68.506	72.982
#1	-.00017	.00098	.00051	.02894	-.09438	.00132	.11194	.00086
#2	-.00017	.00026	.00082	.03073	-.13489	.00174	.05264	.00161
#3	-.00014	.00084	.00088	.05108	-.00155	.00131	.02672	.00028
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00019**	**.03492**	**.00034**	**.00729**	**.00159**	**-.00194**	**.00113**	**.00052**
Stddev	.00031	.01196	.00046	.00336	.00147	.00263	.00402	.00268
%RSD	159.07	34.258	134.06	46.125	92.547	135.75	357.58	520.37
#1	.00038	.02883	.00039	.00624	.00208	.00024	-.00125	.00071
#2	.00036	.04870	-.00014	.00458	.00275	-.00120	-.00114	.00309
#3	-.00016	.02722	.00078	.01105	-.00006	-.00487	.00577	-.00226
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: PBW LJ Acquired: 6/25/2016 22:58:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00001**	**-.00008**	**.00508**	**-.00273**	**.00114**	**.00044**	**.00066**
Stddev	.00164	.00051	.01008	.00394	.00109	.00015	.00112
%RSD	14590.	648.58	198.46	144.23	96.326	34.462	168.97
#1	.00174	.00051	.01609	.00176	.00157	.00053	.00180
#2	-.00150	-.00033	-.00371	-.00561	-.00011	.00026	-.00044
#3	-.00028	-.00041	.00287	-.00433	.00195	.00052	.00063
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10135.**	**62589.**	**3095.8**
Stddev	24.	186.	5.7
%RSD	.23247	.29671	.18319
#1	10143.	62746.	3101.4
#2	10152.	62637.	3095.9
#3	10108.	62384.	3090.1

Approved: June 27, 2016

Sample Name: LCSW LJ Acquired: 6/25/2016 23:02:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19432**	**4.8961**	**.19607**	**.95390**	**.50477**	**.02378**	**5.0869**	**.02467**
Stddev	.00176	.0169	.00036	.00515	.00157	.00015	.0469	.00006
%RSD	.90778	.34589	.18422	.54033	.31152	.61673	.92299	.23915
#1	.19242	4.8827	.19640	.95171	.50641	.02365	5.1361	.02462
#2	.19591	4.9152	.19569	.95978	.50463	.02394	5.0425	.02466
#3	.19461	4.8905	.19612	.95020	.50328	.02375	5.0822	.02474
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09889**	**.24063**	**.25385**	**1.9878**	**24.436**	**.51652**	**4.7044**	**.24710**
Stddev	.00023	.00063	.00116	.0095	.067	.00203	.0278	.00275
%RSD	.23212	.26116	.45751	.47691	.27355	.39333	.59179	1.1124
#1	.09867	.24099	.25369	1.9863	24.386	.51588	4.7289	.24429
#2	.09913	.24100	.25509	1.9791	24.512	.51880	4.6741	.24978
#3	.09888	.23991	.25278	1.9979	24.411	.51489	4.7100	.24724
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51030**	**25.148**	**.24989**	**4.7689**	**.24945**	**.59139**	**.18766**	**2.4779**
Stddev	.00067	.056	.00070	.0091	.00347	.00400	.01211	.0014
%RSD	.13092	.22440	.27941	.19073	1.3920	.67662	6.4536	.05441
#1	.50970	25.212	.24981	4.7624	.25341	.58680	.17389	2.4770
#2	.51102	25.108	.25063	4.7793	.24803	.59411	.19668	2.4795
#3	.51019	25.124	.24924	4.7649	.24691	.59326	.19239	2.4773
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LCSW LJ Acquired: 6/25/2016 23:02:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49047**	**.51082**	**.50961**	**.24925**	**.49196**	**.48864**	**.01928**
Stddev	.00103	.00205	.00391	.00102	.00174	.00019	.00329
%RSD	.20978	.40119	.76774	.40835	.35306	.03834	17.088
#1	.49156	.51042	.51051	.24964	.49257	.48859	.01592
#2	.49032	.50900	.50533	.25000	.49331	.48848	.01940
#3	.48952	.51304	.51300	.24809	.49000	.48885	.02251
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9771.5**	**60209.**	**3050.2**
Stddev	3.0	103.	25.9
%RSD	.03108	.17103	.84982
#1	9769.0	60280.	3044.5
#2	9774.9	60091.	3027.6
#3	9770.6	60258.	3078.5

Approved: June 27, 2016

Sample Name: L1606076404 Acquired: 6/25/2016 23:06:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00050	.02745	.00034	.00861	.00996	.00009	36.605	.00012
Stddev	.00060	.00628	.00209	.00113	.00041	.00005	.105	.00023
%RSD	119.14	22.865	609.72	13.086	4.1637	53.570	.28645	187.24
#1	.00015	.03394	.00014	.00762	.01044	.00014	36.606	.00007
#2	.00016	.02699	-.00164	.00984	.00971	.00010	36.500	.00038
#3	.00120	.02142	.00253	.00837	.00974	.00004	36.710	-.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00009	.01147	.00175	.46715	4.7234	.00625	13.388	.09951
Stddev	.00015	.00054	.00191	.01948	.0341	.00248	.058	.00203
%RSD	171.91	4.6728	109.26	4.1703	.72246	39.751	.43505	2.0444
#1	.00001	.01158	-.00043	.45694	4.6849	.00912	13.373	.10058
#2	-.00027	.01194	.00254	.48961	4.7498	.00477	13.339	.09716
#3	-.00001	.01088	.00314	.45489	4.7356	.00486	13.453	.10078
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00119	34.643	.00396	.03562	.00752	.00339	.00429	4.0453
Stddev	.00028	.078	.00076	.00870	.00150	.00394	.00341	.0166
%RSD	23.933	.22458	19.114	24.436	19.891	116.11	79.423	.40956
#1	.00087	34.732	.00311	.02910	.00894	.00791	.00692	4.0642
#2	.00131	34.587	.00454	.03225	.00766	.00162	.00551	4.0386
#3	.00140	34.610	.00424	.04550	.00596	.00065	.00044	4.0331
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606076404 Acquired: 6/25/2016 23:06:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00088	.19364	.00117	-.00333	.00101	.00286	.00160
Stddev	.00025	.00044	.00456	.00363	.00019	.00020	.00097
%RSD	28.325	.22814	391.45	108.96	18.806	7.1675	60.613
#1	.00083	.19387	-.00007	-.00660	.00107	.00306	.00048
#2	.00116	.19313	.00622	-.00396	.00116	.00265	.00224
#3	.00066	.19392	-.00265	.00057	.00080	.00286	.00207
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9727.8	59976.	3053.1
Stddev	23.9	148.	4.4
%RSD	.24547	.24636	.14325
#1	9700.3	60146.	3049.4
#2	9741.5	59877.	3057.9
#3	9741.8	59905.	3052.0

Approved: June 27, 2016

Sample Name: L1606076405S Acquired: 6/25/2016 23:10:17 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19582**	**4.9527**	**.19879**	**.95673**	**.51438**	**.02432**	**42.805**	**.02505**
Stddev	.00335	.0137	.00128	.00425	.00065	.00003	.112	.00030
%RSD	1.7110	.27715	.64557	.44412	.12542	.14170	.26049	1.1988
#1	.19856	4.9529	.19764	.95186	.51480	.02435	42.707	.02530
#2	.19683	4.9664	.19856	.95967	.51471	.02433	42.926	.02513
#3	.19208	4.9389	.20017	.95866	.51364	.02429	42.781	.02472
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09757**	**.25598**	**.24917**	**2.4354**	**29.536**	**.52338**	**18.805**	**.34995**
Stddev	.00078	.00042	.00127	.0181	.040	.00393	.025	.00506
%RSD	.79936	.16278	.51136	.74393	.13605	.74997	.13071	1.4469
#1	.09813	.25630	.25060	2.4222	29.582	.51885	18.805	.34467
#2	.09790	.25551	.24817	2.4278	29.516	.52547	18.830	.35476
#3	.09668	.25612	.24874	2.4560	29.509	.52582	18.780	.35044
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51541**	**61.224**	**.25128**	**4.8968**	**.24942**	**.59855**	**.19087**	**6.6449**
Stddev	.00129	.228	.00073	.0088	.00234	.00299	.00293	.0075
%RSD	.25084	.37164	.28907	.17909	.93699	.50010	1.5365	.11257
#1	.51663	61.311	.25055	4.9006	.24863	.59583	.18976	6.6461
#2	.51554	61.394	.25129	4.9031	.24758	.60176	.19420	6.6517
#3	.51406	60.965	.25200	4.8868	.25205	.59805	.18866	6.6369
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606076405S Acquired: 6/25/2016 23:10:17 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.49233	.70070	.50542	.24860	.50000	.48511	.00505
Stddev	.00086	.00402	.00572	.00107	.00163	.00066	.00072
%RSD	.17371	.57387	1.1319	.43103	.32683	.13619	14.263
#1	.49184	.70101	.51171	.24742	.49924	.48572	.00441
#2	.49332	.70456	.50403	.24884	.49889	.48521	.00491
#3	.49184	.69653	.50053	.24953	.50188	.48441	.00583
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9652.6	59187.	3022.5
Stddev	17.6	118.	14.2
%RSD	.18215	.19875	.46825
#1	9645.4	59224.	3018.6
#2	9639.7	59281.	3010.7
#3	9672.6	59055.	3038.2

Approved: June 27, 2016

Sample Name: L1606076406SD Acquired: 6/25/2016 23:13:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19161**	**4.9186**	**.19913**	**.95429**	**.51211**	**.02396**	**42.977**	**.02467**
Stddev	.00161	.0125	.00343	.00448	.00436	.00009	.276	.00017
%RSD	.83957	.25322	1.7203	.46971	.85098	.36242	.64279	.68812
#1	.19148	4.9258	.19753	.95898	.51664	.02396	43.237	.02475
#2	.19328	4.9042	.19680	.95004	.51174	.02387	43.008	.02479
#3	.19007	4.9258	.20306	.95385	.50794	.02404	42.687	.02448
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09706**	**.25178**	**.24789**	**2.4233**	**29.585**	**.51807**	**18.712**	**.35011**
Stddev	.00057	.00068	.00034	.0356	.112	.00327	.182	.00396
%RSD	.58951	.26961	.13796	1.4688	.37990	.63166	.97051	1.1325
#1	.09772	.25103	.24806	2.4608	29.711	.51534	18.898	.35283
#2	.09684	.25235	.24811	2.4189	29.497	.52170	18.704	.34556
#3	.09664	.25196	.24749	2.3900	29.545	.51719	18.535	.35193
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51262**	**61.425**	**.24874**	**4.8787**	**.24757**	**.59558**	**.18865**	**6.6917**
Stddev	.00159	.310	.00133	.0141	.00184	.00348	.00465	.0131
%RSD	.30991	.50524	.53301	.28958	.74220	.58390	2.4654	.19574
#1	.51304	61.773	.24994	4.8746	.24617	.59720	.19309	6.6816
#2	.51396	61.322	.24897	4.8944	.24965	.59795	.18905	6.7065
#3	.51087	61.178	.24732	4.8670	.24689	.59159	.18381	6.6870
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606076406SD Acquired: 6/25/2016 23:13:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.48585	.69992	.50428	.08491	.49039	.49499	.00464
Stddev	.00076	.00443	.00835	.00108	.00225	.00372	.00157
%RSD	.15628	.63319	1.6562	1.2751	.45854	.75207	33.818
#1	.48649	.70500	.51389	.08368	.49168	.49197	.00317
#2	.48606	.69789	.49873	.08571	.48779	.49386	.00446
#3	.48501	.69687	.50023	.08535	.49169	.49915	.00630
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9674.4	59675.	3057.4
Stddev	5.6	128.	11.6
%RSD	.05826	.21445	.38044
#1	9670.2	59575.	3048.3
#2	9680.8	59819.	3053.4
#3	9672.1	59630.	3070.5

Approved: June 27, 2016

Sample Name: L1606099102 Acquired: 6/25/2016 23:17:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 10 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00035	.08527	.00559	.04109	.01538	.00019	.57467	.00159
Stddev	.00233	.00869	.00189	.00114	.00133	.00004	.01094	.00013
%RSD	663.84	10.192	33.720	2.7694	8.6207	22.830	1.9043	7.9131
#1	.00210	.09496	.00628	.04240	.01605	.00016	.56261	.00167
#2	-.00061	.08272	.00346	.04040	.01386	.00017	.57746	.00145
#3	-.00255	.07815	.00704	.04046	.01624	.00024	.58395	.00167
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00096	.05412	.00582	.38755	1.2527	.00289	.47003	.00071
Stddev	.00015	.00024	.00026	.01093	.0635	.00116	.03189	.00162
%RSD	15.476	.43524	4.4250	2.8204	5.0656	40.155	6.7852	226.02
#1	.00113	.05437	.00590	.40001	1.1922	.00164	.49884	-.00098
#2	.00091	.05406	.00602	.38303	1.3188	.00308	.43576	.00089
#3	.00085	.05391	.00553	.37960	1.2473	.00394	.47550	.00224
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00360	218.01	.00389	.36041	.01046	-.00372	.00482	1.3728
Stddev	.00017	.12	.00068	.00488	.00162	.00083	.00193	.0038
%RSD	4.7149	.05666	17.447	1.3551	15.457	22.416	40.114	.27490
#1	.00375	218.03	.00311	.35572	.01225	-.00460	.00284	1.3685
#2	.00365	218.13	.00423	.36547	.01003	-.00293	.00492	1.3754
#3	.00341	217.88	.00434	.36004	.00911	-.00364	.00670	1.3745
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606099102 Acquired: 6/25/2016 23:17:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 10 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00632**	**.09765**	**.07700**	**-.00067**	**.01524**	**.07845**	**.01879**
Stddev	.00056	.00044	.00219	.00203	.00068	.00014	.00089
%RSD	8.9139	.45291	2.8381	300.30	4.4598	.18118	4.7496
#1	.00609	.09718	.07562	.00166	.01602	.07833	.01918
#2	.00590	.09770	.07952	-.00172	.01485	.07841	.01777
#3	.00696	.09806	.07586	-.00197	.01484	.07861	.01943
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10241.**	**62264.**	**3185.5**
Stddev	15.	196.	4.7
%RSD	.14516	.31405	.14616
#1	10253.	62395.	3189.9
#2	10245.	62357.	3185.8
#3	10224.	62039.	3180.6

Approved: June 27, 2016

Sample Name: L1606099102PS Acquired: 6/25/2016 23:21:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 10 Custom ID2: Custom ID3:
Comment: WG573952-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.20426**	**5.2585**	**.21569**	**1.0545**	**.53374**	**.02552**	**5.8510**	**.02789**
Stddev	.00128	.0258	.00476	.0057	.01070	.00022	.0799	.00005
%RSD	.62816	.49053	2.2048	.54061	2.0041	.85107	1.3656	.17945
#1	.20288	5.2399	.21984	1.0489	.54587	.02533	5.9343	.02786
#2	.20446	5.2478	.21673	1.0543	.52973	.02547	5.7750	.02786
#3	.20543	5.2880	.21050	1.0603	.52563	.02576	5.8438	.02795
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.10258**	**.30280**	**.26705**	**2.4069**	**27.035**	**.52853**	**5.3059**	**.25725**
Stddev	.00041	.00101	.00181	.0614	.546	.00889	.1709	.00709
%RSD	.40425	.33448	.67841	2.5526	2.0183	1.6821	3.2205	2.7577
#1	.10253	.30184	.26742	2.4708	27.658	.53822	5.4655	.26479
#2	.10301	.30272	.26864	2.4018	26.642	.52076	5.1256	.25071
#3	.10219	.30386	.26508	2.3482	26.805	.52661	5.3266	.25626
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.52291**	**252.79**	**.26034**	**5.5827**	**.26863**	**.62508**	**.20920**	**4.0596**
Stddev	.00363	5.13	.00069	.0240	.00232	.00165	.00657	.0117
%RSD	.69330	2.0298	.26564	.43012	.86233	.26363	3.1385	.28905
#1	.52662	258.69	.26066	5.6009	.26650	.62318	.21182	4.0655
#2	.52272	250.27	.26082	5.5917	.26829	.62599	.21406	4.0673
#3	.51938	249.40	.25955	5.5555	.27109	.62608	.20173	4.0461
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606099102PS Acquired: 6/25/2016 23:21:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 10 Custom ID2: Custom ID3:
Comment: WG573952-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51599**	**.62613**	**.61043**	**.25381**	**.52664**	**.60277**	**.02286**
Stddev	.00198	.01306	.01573	.00527	.00286	.00158	.00057
%RSD	.38295	2.0851	2.5770	2.0756	.54391	.26285	2.4777
#1	.51752	.64119	.62824	.25887	.52456	.60412	.02234
#2	.51668	.61903	.59844	.25420	.52545	.60318	.02277
#3	.51376	.61815	.60462	.24835	.52991	.60103	.02347
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9624.7**	**58159.**	**2980.0**
Stddev	48.4	89.	45.7
%RSD	.50312	.15320	1.5350
#1	9582.5	58170.	2928.1
#2	9614.0	58242.	2997.3
#3	9677.6	58064.	3014.5

Approved: June 27, 2016

Sample Name: L1606099102SDL Acquired: 6/25/2016 23:25:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 50 Custom ID2: Custom ID3:
Comment: WG573952-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00117	.01283	.00236	.01119	.00550	.00018	.08026	.00026
Stddev	.00047	.00301	.00204	.00193	.00071	.00008	.03574	.00006
%RSD	40.405	23.474	86.372	17.221	12.910	46.045	44.534	24.105
#1	.00083	.01247	.00021	.01333	.00520	.00027	.04235	.00026
#2	.00098	.01601	.00261	.01064	.00630	.00011	.11335	.00020
#3	.00172	.01002	.00427	.00960	.00498	.00015	.08510	.00032
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00022	.01258	.00098	.10105	.18990	-.00151	.06238	.00028
Stddev	.00006	.00091	.00048	.00767	.08103	.00626	.02555	.00271
%RSD	25.732	7.2135	49.236	7.5891	42.670	413.92	40.953	970.07
#1	.00026	.01275	.00118	.09700	.21903	.00544	.04558	.00279
#2	.00015	.01339	.00043	.10990	.09833	-.00328	.04979	.00064
#3	.00024	.01160	.00132	.09626	.25235	-.00669	.09179	-.00259
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00205	47.703	.00017	.08239	.00525	-.00365	.00219	.31315
Stddev	.00019	.192	.00051	.00485	.00385	.00400	.00870	.00335
%RSD	9.3003	.40144	294.11	5.8889	73.348	109.43	397.40	1.0688
#1	.00183	47.834	.00076	.07694	.00083	-.00754	.00399	.31041
#2	.00217	47.483	-.00009	.08400	.00786	-.00387	-.00727	.31216
#3	.00215	47.792	-.00015	.08624	.00708	.00045	.00985	.31688
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606099102SDL Acquired: 6/25/2016 23:25:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 50 Custom ID2: Custom ID3:
Comment: WG573952-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00128**	**.02176**	**.02063**	**.00132**	**.00408**	**.01800**	**.00415**
Stddev	.00053	.00039	.00638	.00109	.00052	.00037	.00045
%RSD	41.627	1.7773	30.936	82.221	12.786	2.0724	10.786
#1	.00158	.02176	.02171	.00142	.00446	.01762	.00425
#2	.00067	.02138	.02640	.00236	.00348	.01802	.00454
#3	.00160	.02215	.01377	.00019	.00429	.01836	.00366
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9804.5**	**59680.**	**2983.6**
Stddev	22.5	103.	15.3
%RSD	.22901	.17203	.51306
#1	9824.1	59682.	2991.0
#2	9809.6	59782.	2993.9
#3	9780.0	59577.	2966.0

Approved: June 27, 2016

Sample Name: L1606099501 Acquired: 6/25/2016 23:29:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 1000 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00086	.00231	.00190	.02254	.12371	.00015	15.528	-.00008
Stddev	.00014	.00712	.00142	.00137	.00163	.00003	.200	.00022
%RSD	16.093	308.09	74.696	6.0976	1.3212	21.126	1.2862	282.14
#1	.00100	.01032	.00271	.02232	.12196	.00014	15.307	.00002
#2	.00072	-.00013	.00272	.02129	.12520	.00012	15.695	.00008
#3	.00085	-.00326	.00026	.02401	.12396	.00018	15.583	-.00034
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00031	.00027	.00136	.08102	.36083	.08036	1.7207	.01281
Stddev	.00017	.00026	.00081	.01723	.08162	.00488	.0426	.00381
%RSD	55.449	94.239	59.356	21.269	22.620	6.0696	2.4776	29.711
#1	.00018	.00004	.00043	.07010	.45158	.07724	1.7124	.01021
#2	.00026	.00024	.00186	.07209	.29342	.07785	1.6828	.01718
#3	.00051	.00055	.00179	.10089	.33750	.08598	1.7669	.01103
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00006	40.169	.00020	.00739	.00147	-.00261	.00048	.01475
Stddev	.00007	.358	.00037	.00186	.00129	.00197	.00359	.00082
%RSD	116.52	.89035	187.93	25.143	87.719	75.441	741.63	5.5637
#1	.00010	39.803	.00022	.00902	.00206	-.00263	-.00044	.01382
#2	.00008	40.518	.00056	.00537	.00236	-.00063	-.00256	.01539
#3	-.00002	40.186	-.00019	.00779	-.00001	-.00456	.00445	.01503
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606099501 Acquired: 6/25/2016 23:29:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 1000 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	2.0444	.00152	-.00059	-.00017	.00315	.00110
Stddev	.00034	.0207	.00598	.00278	.00059	.00019	.00107
%RSD	58.480	1.0109	393.49	469.74	343.26	6.1207	97.322
#1	.00020	2.0228	.00742	-.00167	.00024	.00305	.00162
#2	.00070	2.0639	.00167	-.00267	-.00085	.00303	-.00013
#3	.00086	2.0465	-.00453	.00256	.00009	.00337	.00180
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9773.4	59741.	3000.7
Stddev	10.5	113.	31.9
%RSD	.10719	.18996	1.0628
#1	9785.4	59856.	3037.0
#2	9766.5	59737.	2977.2
#3	9768.1	59629.	2987.9

Approved: June 27, 2016

Sample Name: L1606099502 Acquired: 6/25/2016 23:32:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 1000 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00001	-.00479	-.00049	.02240	.12357	.00014	15.278	-.00007
Stddev	.00142	.00423	.00253	.00505	.00148	.00006	.144	.00014
%RSD	22192.	88.307	518.91	22.533	1.1984	44.965	.94397	206.25
#1	.00014	-.00965	-.00218	.02478	.12272	.00011	15.115	-.00014
#2	.00133	-.00281	-.00170	.02581	.12271	.00021	15.329	.00010
#3	-.00149	-.00191	.00242	.01660	.12528	.00010	15.390	-.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	.00065	.00042	.06837	.37746	.07803	1.6184	.01419
Stddev	.00041	.00112	.00027	.00656	.03147	.00401	.0513	.00310
%RSD	83.614	171.25	62.991	9.6001	8.3371	5.1380	3.1674	21.849
#1	.00050	-.00064	.00063	.07323	.35510	.08210	1.6690	.01383
#2	.00090	.00123	.00012	.06091	.36384	.07791	1.6197	.01745
#3	.00007	.00137	.00052	.07098	.41345	.07408	1.5665	.01128
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00036	39.920	-.00037	.00510	.00199	-.00064	.00335	.01167
Stddev	.00036	.366	.00054	.01025	.00130	.00317	.00244	.00126
%RSD	102.18	.91613	144.37	200.85	65.670	494.36	72.931	10.832
#1	-.00077	39.520	-.00039	.00961	.00071	-.00417	.00384	.01302
#2	-.00018	40.239	-.00090	.01233	.00193	.00029	.00550	.01150
#3	-.00011	40.000	.00017	-.00663	.00331	.00196	.00070	.01051
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606099502 Acquired: 6/25/2016 23:32:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 1000 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00055**	**2.0217**	**-.00337**	**-.00055**	**.00074**	**.00315**	**.00087**
Stddev	.00116	.0191	.00559	.00170	.00084	.00020	.00103
%RSD	210.74	.94683	165.64	307.06	114.28	6.4714	118.16
#1	-.00142	2.0010	.00194	-.00241	.00129	.00330	.00185
#2	.00077	2.0388	-.00920	-.00018	-.00023	.00292	.00097
#3	-.00100	2.0253	-.00287	.00093	.00117	.00324	-.00020
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9730.5**	**59619.**	**2989.3**
Stddev	18.2	229.	29.6
%RSD	.18733	.38360	.99126
#1	9750.6	59882.	3021.9
#2	9715.1	59468.	2964.1
#3	9725.8	59506.	2981.9

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 23:36:47 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40930**	**10.238**	**.40819**	**.50228**	**1.0459**	**.04973**	**10.404**
Stddev	.00043	.021	.00043	.00341	.0044	.00005	.020
%RSD	.10460	.20815	.10461	.67892	.41689	.09834	.19571
#1	.40883	10.251	.40851	.50622	1.0496	.04974	10.381
#2	.40944	10.250	.40771	.50027	1.0411	.04968	10.418
#3	.40965	10.214	.40835	.50036	1.0470	.04978	10.414
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05210**	**.20379**	**.49816**	**.52303**	**4.0622**	**51.067**	**1.0469**
Stddev	.00019	.00042	.00279	.00142	.0372	.160	.0078
%RSD	.35747	.20724	.56062	.27079	.91500	.31299	.74725
#1	.05231	.20342	.50024	.52229	4.1047	51.143	1.0555
#2	.05194	.20371	.49926	.52467	4.0462	50.883	1.0402
#3	.05207	.20425	.49499	.52215	4.0357	51.175	1.0449
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.7214**	**.51746**	**1.0407**	**52.052**	**.51652**	**10.202**	**.51928**
Stddev	.0660	.00152	.0096	.278	.00085	.015	.00145
%RSD	.67863	.29343	.92144	.53407	.16386	.14443	.27838
#1	9.7320	.51900	1.0479	52.343	.51582	10.200	.51894
#2	9.7814	.51597	1.0444	51.789	.51627	10.218	.52086
#3	9.6508	.51742	1.0298	52.022	.51746	10.189	.51803
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 23:36:47 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2277**	**.40544**	**5.2555**	**1.0220**	**1.0464**	**1.0514**	**.52265**
Stddev	.0051	.00303	.0032	.0024	.0054	.0078	.00372
%RSD	.41461	.74796	.06070	.23053	.51974	.73901	.71099
#1	1.2332	.40616	5.2590	1.0221	1.0521	1.0443	.52536
#2	1.2232	.40805	5.2526	1.0196	1.0413	1.0597	.52418
#3	1.2265	.40211	5.2550	1.0243	1.0458	1.0502	.51841
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0117**	**1.0320**	F **2.9933**
Stddev	.0025	.0009	.0109
%RSD	.24458	.08681	.36389
#1	1.0144	1.0330	3.0053
#2	1.0111	1.0312	2.9841
#3	1.0096	1.0318	2.9905
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9691.0**	**59432.**	**3012.0**
Stddev	5.1	44.	12.8
%RSD	.05245	.07397	.42554
#1	9696.2	59404.	2998.2
#2	9686.1	59410.	3023.6
#3	9690.7	59483.	3014.1

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 23:40:25 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00145**	**-.00660**	**-.00143**	**.00292**	**.00177**	**.00016**	**-.01784**	**.00008**
Stddev	.00180	.00657	.00282	.00130	.00082	.00003	.03649	.00007
%RSD	124.44	99.487	197.72	44.383	46.493	20.281	204.59	79.919
#1	.00061	-.01208	-.00468	.00217	.00185	.00015	-.05680	.00016
#2	-.00276	-.00840	.00028	.00442	.00255	.00019	-.01224	.00005
#3	-.00219	.00068	.00012	.00218	.00091	.00013	.01553	.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**.00038**	**.00052**	**.02154**	**-.09198**	**-.00110**	**-.01322**	**.00018**
Stddev	.00024	.00088	.00138	.01483	.10535	.00141	.06861	.00246
%RSD	523.34	234.01	265.29	68.833	114.54	127.62	519.13	1371.8
#1	-.00015	.00090	-.00012	.01936	.02327	-.00076	.06137	-.00117
#2	-.00022	-.00064	-.00042	.03733	-.18333	-.00265	-.02738	.00302
#3	.00023	.00086	.00210	.00792	-.11589	.00010	-.07365	-.00132
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00348**	**.04821**	**.00016**	**.00585**	**.00443**	**.00314**	**.00156**	**.03553**
Stddev	.00065	.02773	.00095	.00367	.00371	.00275	.00114	.00188
%RSD	18.560	57.517	582.95	62.725	83.852	87.739	73.200	5.2906
#1	.00289	.07448	.00125	.00268	.00440	.00563	.00189	.03713
#2	.00338	.05093	-.00027	.00499	.00073	.00018	.00029	.03346
#3	.00417	.01922	-.00050	.00987	.00816	.00360	.00249	.03599
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 23:40:25 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00025**	**.00051**	**.00604**	**-.00528**	**.00062**	**.00012**	**.00482**
Stddev	.00070	.00006	.00458	.00163	.00029	.00008	.00094
%RSD	280.85	11.978	75.793	30.927	46.562	68.537	19.510
#1	.00047	.00050	.00150	-.00357	.00036	.00006	.00376
#2	.00081	.00057	.01066	-.00682	.00057	.00008	.00556
#3	-.00053	.00045	.00597	-.00546	.00093	.00021	.00513
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9759.5**	**60265.**	**3013.5**
Stddev	10.8	85.	17.4
%RSD	.11040	.14047	.57599
#1	9754.6	60307.	3032.0
#2	9771.9	60320.	2997.5
#3	9752.1	60167.	3011.1

Approved: June 27, 2016

Sample Name: ICSA Acquired: 6/25/2016 23:44:20 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F -.00562	263.10	.00178	-.00976	.00140	.00014	248.10
Stddev	.00117	.16	.00383	.00309	.00113	.00004	.61
%RSD	20.823	.06041	215.50	31.614	80.449	26.067	.24782
#1	-.00542	263.25	-.00154	-.01064	.00266	.00012	248.67
#2	-.00456	262.94	.00597	-.00633	.00047	.00013	247.45
#3	-.00688	263.13	.00090	-.01231	.00109	.00019	248.17
Check ?	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit	.00400						
Low Limit	-.00400						

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00017	-.00060	-.00056	F .01039	95.630	-.04599	.01242
Stddev	.00016	.00022	.00039	.00101	.248	.03226	.00307
%RSD	92.869	36.608	70.420	9.7281	.25928	70.136	24.681
#1	-.00031	-.00080	-.00098	.00928	95.904	-.06101	.01295
#2	-.00020	-.00063	-.00021	.01127	95.420	-.06800	.01519
#3	.00000	-.00037	-.00048	.01062	95.566	-.00896	.00913
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				.00400			
Low Limit				-.00400			

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	236.10	.00050	-.00055	.08182	-.00122	.03425	.00395
Stddev	.41	.00175	.00028	.01035	.00129	.00383	.00558
%RSD	.17524	347.68	51.671	12.652	105.94	11.172	141.13
#1	236.13	.00226	-.00079	.09323	-.00097	.03151	-.00194
#2	235.67	.00051	-.00062	.07920	-.00262	.03262	.00465
#3	236.50	-.00125	-.00024	.07303	-.00007	.03862	.00916
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: ICSA Acquired: 6/25/2016 23:44:20 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00072	-.00782	.00827	.00013	.00094	-.00929	-.00404
Stddev	.00319	.00204	.00198	.00005	.00018	.00784	.00342
%RSD	442.40	26.134	23.931	35.338	18.766	84.463	84.650
#1	.00212	-.01015	.01018	.00009	.00088	-.00188	-.00347
#2	-.00417	-.00696	.00842	.00012	.00080	-.01750	-.00771
#3	-.00012	-.00634	.00622	.00018	.00114	-.00848	-.00094
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	-.00163	-.00560	-.01351
Stddev	.00058	.00018	.00078
%RSD	35.475	3.3032	5.7782
#1	-.00221	-.00565	-.01421
#2	-.00162	-.00574	-.01366
#3	-.00105	-.00539	-.01267
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9256.6	56705.	2976.7
Stddev	13.6	91.	12.8
%RSD	.14727	.16098	.43054
#1	9241.1	56650.	2963.0
#2	9266.9	56811.	2988.4
#3	9261.8	56655.	2978.7

Approved: June 27, 2016

Sample Name: ICSAB Acquired: 6/25/2016 23:48:09 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.52011**	**271.09**	**.26014**	**-.01193**	**.26131**	**.24391**	**254.06**	**.49893**
Stddev	.00309	.31	.00108	.00072	.00207	.00022	1.57	.00035
%RSD	.59470	.11617	.41599	6.0451	.79127	.08930	.61634	.06930
#1	.51910	271.29	.26138	-.01172	.26259	.24374	255.54	.49856
#2	.51765	270.73	.25938	-.01274	.26241	.24383	254.23	.49898
#3	.52358	271.26	.25965	-.01134	.25892	.24415	252.42	.49925
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.24094**	**.25041**	**.26416**	**96.077**	**5.2095**	**.01124**	**238.55**	**.24627**
Stddev	.00056	.00142	.00058	.420	.1352	.00179	1.07	.00329
%RSD	.23299	.56520	.21819	.43687	2.5957	15.928	.44935	1.3378
#1	.24046	.25065	.26387	96.381	5.3478	.01132	239.59	.24346
#2	.24081	.24889	.26378	96.252	5.2032	.00941	238.63	.24545
#3	.24156	.25169	.26482	95.598	5.0776	.01299	237.45	.24990
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00041**	**5.4959**	**.48742**	**.09841**	**.49766**	**.51195**	**.23776**	**.00256**
Stddev	.00052	.0364	.00161	.00635	.00144	.00162	.00501	.00193
%RSD	128.92	.66237	.33105	6.4479	.28893	.31661	2.1076	75.415
#1	-.00068	5.5313	.48557	.09364	.49656	.51344	.24335	.00339
#2	-.00073	5.4979	.48847	.09597	.49929	.51023	.23366	.00393
#3	.00020	5.4586	.48823	.10561	.49712	.51219	.23626	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: ICSAB Acquired: 6/25/2016 23:48:09 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00140**	**.00103**	**-.00323**	**.47273**	**.25455**	**.48271**	**-.01305**
Stddev	.00059	.00027	.00250	.00229	.00097	.00045	.00046
%RSD	41.750	26.302	77.431	.48347	.38190	.09250	3.5206
#1	.00144	.00120	-.00356	.47420	.25352	.48220	-.01343
#2	.00197	.00072	-.00058	.47010	.25467	.48293	-.01317
#3	.00080	.00116	-.00555	.47390	.25546	.48302	-.01254
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9240.5**	**56174.**	**2955.7**
Stddev	9.6	118.	17.7
%RSD	.10383	.20921	.59797
#1	9251.6	56039.	2946.0
#2	9234.1	56253.	2944.9
#3	9235.9	56229.	2976.1

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 23:51:52 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.41108	10.316	.40843	.50327	1.0436	.05024	10.459
Stddev	.00161	.013	.00298	.00077	.0011	.00011	.030
%RSD	.39215	.12124	.73007	.15382	.10203	.22382	.28694
#1	.41119	10.324	.40804	.50264	1.0446	.05015	10.464
#2	.41263	10.302	.40566	.50304	1.0438	.05037	10.487
#3	.40941	10.323	.41158	.50413	1.0425	.05022	10.428
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05234	.20485	.50201	.52446	4.0759	51.088	1.0484
Stddev	.00020	.00025	.00165	.00123	.0669	.060	.0059
%RSD	.38699	.12298	.32889	.23505	1.6409	.11683	.55819
#1	.05251	.20498	.50023	.52357	4.1126	51.031	1.0491
#2	.05240	.20456	.50349	.52586	3.9987	51.083	1.0423
#3	.05212	.20501	.50230	.52393	4.1164	51.150	1.0539
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.8126	.50786	1.0468	51.847	.51713	10.210	.51778
Stddev	.0274	.00264	.0020	.030	.00045	.013	.00098
%RSD	.27960	.52030	.19073	.05848	.08757	.13121	.18906
#1	9.8005	.51090	1.0485	51.876	.51722	10.201	.51858
#2	9.7933	.50648	1.0474	51.816	.51663	10.204	.51807
#3	9.8440	.50619	1.0446	51.848	.51752	10.226	.51669
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 23:51:52 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2406**	**.41201**	**5.2627**	**1.0255**	**1.0428**	**1.0401**	**.52263**
Stddev	.0022	.00564	.0065	.0018	.0016	.0094	.00276
%RSD	.17501	1.3693	.12412	.17227	.15774	.90409	.52737
#1	1.2431	.41191	5.2553	1.0271	1.0421	1.0459	.52461
#2	1.2392	.40642	5.2651	1.0236	1.0417	1.0292	.52380
#3	1.2396	.41770	5.2677	1.0258	1.0447	1.0451	.51948
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0216**	**1.0330**	F **3.0271**
Stddev	.0027	.0013	.0163
%RSD	.26838	.12831	.53748
#1	1.0227	1.0315	3.0446
#2	1.0236	1.0335	3.0125
#3	1.0185	1.0340	3.0242
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9680.4**	**59250.**	**3000.1**
Stddev	28.5	27.	16.9
%RSD	.29390	.04497	.56388
#1	9711.9	59280.	2982.6
#2	9672.8	59242.	3016.3
#3	9656.6	59229.	3001.5

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 23:55:30 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00125	.00182	.00085	.00036	.00255	.00017	-.02838	-.00012
Stddev	.00082	.00761	.00304	.00225	.00150	.00005	.02296	.00019
%RSD	65.454	418.99	358.52	623.67	58.837	28.075	80.890	157.58
#1	.00189	-.00688	.00358	-.00006	.00362	.00012	-.00325	-.00021
#2	.00153	.00513	.00139	-.00165	.00321	.00020	-.04827	-.00024
#3	.00033	.00720	-.00242	.00279	.00083	.00021	-.03363	.00010
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.00075	.00113	.02762	-.02056	.00086	-.01380	-.00101
Stddev	.00025	.00070	.00060	.00527	.01321	.00185	.01535	.00114
%RSD	279.31	93.710	52.614	19.071	64.270	216.68	111.21	112.81
#1	-.00016	.00013	.00054	.02156	-.02746	-.00001	-.01846	.00029
#2	.00010	.00151	.00112	.03019	-.02889	-.00040	-.02627	-.00181
#3	.00033	.00061	.00173	.03110	-.00532	.00298	.00334	-.00151
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00326	.05197	-.00006	.00537	-.00040	.00432	.00298	.03440
Stddev	.00030	.03661	.00066	.00611	.00131	.00120	.00482	.00043
%RSD	9.1019	70.433	1169.2	113.80	328.70	27.704	161.98	1.2407
#1	.00299	.03235	.00066	.00591	.00062	.00425	-.00005	.03422
#2	.00322	.09421	-.00018	.01119	-.00187	.00556	.00044	.03489
#3	.00358	.02936	-.00065	-.00099	.00005	.00317	.00854	.03409
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 23:55:30 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**.00044**	**.00562**	**-.00278**	**.00100**	**.00010**	**.00427**
Stddev	.00044	.00041	.00677	.00064	.00081	.00020	.00057
%RSD	558.27	93.814	120.49	22.928	80.799	205.25	13.474
#1	.00052	-.00002	.00467	-.00264	.00112	.00024	.00360
#2	.00007	.00078	.01282	-.00348	.00014	-.00013	.00461
#3	-.00036	.00055	-.00063	-.00223	.00174	.00017	.00459
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9746.4**	**60216.**	**3006.7**
Stddev	3.8	76.	7.9
%RSD	.03876	.12554	.26429
#1	9742.9	60301.	3002.9
#2	9746.0	60157.	3001.3
#3	9750.4	60189.	3015.8

Approved: June 27, 2016

Sample Name: LLCCV Acquired: 6/25/2016 23:59:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00916	.16469	.00814	.08024	.01010	.00165	.38196	.00067
Stddev	.00085	.00884	.00209	.00058	.00103	.00003	.03578	.00008
%RSD	9.3027	5.3671	25.632	.71975	10.222	1.6058	9.3673	12.378
#1	.00982	.16059	.01037	.08090	.01024	.00166	.39925	.00061
#2	.00944	.15864	.00624	.07981	.01105	.00162	.40581	.00076
#3	.00820	.17483	.00780	.08002	.00900	.00168	.34082	.00063
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00411	.00454	.00520	.10552	.68660	.08825	.42307	.00800
Stddev	.00033	.00070	.00174	.01827	.09614	.00099	.09575	.00252
%RSD	7.9024	15.364	33.422	17.313	14.002	1.1218	22.633	31.490
#1	.00383	.00377	.00720	.10444	.59286	.08799	.31358	.00534
#2	.00447	.00514	.00437	.12431	.68197	.08934	.49110	.01035
#3	.00404	.00471	.00403	.08782	.78497	.08741	.46454	.00831
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00870	.45736	.01598	.79373	.00972	.08303	.01581	.78238
Stddev	.00060	.01151	.00062	.00779	.00188	.00162	.00507	.00288
%RSD	6.8723	2.5158	3.9084	.98133	19.382	1.9461	32.090	.36823
#1	.00805	.45897	.01600	.80272	.00902	.08401	.02162	.78192
#2	.00924	.44514	.01534	.78909	.01186	.08117	.01226	.78546
#3	.00881	.46798	.01659	.78938	.00829	.08392	.01355	.77975
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LLCCV Acquired: 6/25/2016 23:59:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40813**	**.04261**	**.02883**	**.15405**	**.00896**	**.01738**	**.13453**
Stddev	.00156	.00009	.00723	.00166	.00036	.00005	.00136
%RSD	.38342	.20072	25.085	1.0771	4.0551	.28767	1.0139
#1	.40866	.04260	.02755	.15254	.00855	.01741	.13450
#2	.40935	.04253	.02233	.15583	.00907	.01742	.13318
#3	.40636	.04270	.03662	.15377	.00925	.01733	.13591
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9793.5**	**60446.**	**3011.5**
Stddev	14.7	195.	8.9
%RSD	.15029	.32191	.29401
#1	9810.3	60377.	3006.1
#2	9783.4	60295.	3006.7
#3	9786.6	60666.	3021.7

Approved: June 27, 2016

2.3.2 Metals ICP-MS Data

2.3.2.1 Summary Data

Microbac

Lab Report #: L16060822
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060822-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18WW03F-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/28/2016 13:15
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/28/2016 20:42
Collect Date:	06/14/2016 07:40	**Dilution:**	1	**File ID:**	NI.062816.204236
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.000919	J	0.00200	0.00100	0.000500
Barium, Total	7440-39-3	0.173		0.00600	0.00300	0.00150
Cadmium, Total	7440-43-9	0.000600	U	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00200	U	0.00400	0.00200	0.00100
Cobalt, Total	7440-48-4	0.00100	U	0.00200	0.00100	0.000500
Copper, Total	7440-50-8	0.00200	U	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00100	U	0.00200	0.00100	0.000500
Manganese, Total	7439-96-5	0.0960		0.00400	0.00200	0.00100
Nickel, Total	7440-02-0	0.00400	U	0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.00100	U	0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0250	U	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060822
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060822-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18CPTMW04SW-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/28/2016 13:15
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/28/2016 20:45
Collect Date:	06/14/2016 09:00	**Dilution:**	1	**File ID:**	NI.062816.204547
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00407		0.00200	0.00100	0.000500
Cadmium, Total	7440-43-9	0.000600	U	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00950		0.00400	0.00200	0.00100
Cobalt, Total	7440-48-4	0.00368		0.00200	0.00100	0.000500
Copper, Total	7440-50-8	0.00200	U	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00100	U	0.00200	0.00100	0.000500
Manganese, Total	7439-96-5	0.213		0.00400	0.00200	0.00100
Nickel, Total	7440-02-0	0.00574	J	0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.00100	U	0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0250	U	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060822
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060822-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18CPTMW04SW-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/29/2016 12:25
Collect Date:	06/14/2016 09:00	**Dilution:**	10	**File ID:**	NI.062916.122501
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Barium, Total	7440-39-3	0.936		0.0600	0.0300	0.0150

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060822
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.3.2.2 QC Summary Data

Example 6020 Calculations
Perkin Elmer NexION 300X

1.0 Initial Calibration (ICAL) Parameters
The system performs linear regression from data consisting of a blank and three standards.

2.0 Calculating the concentration (C) of an element in water using data from prep log, run log, and quantitation report (note:the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (ug/L)	0.1
Vf = Final volume	100
Vi = Initial volume	40
D = Dilution factor as a multiplier (10X = 10)	1
Cx = Concentration of element in (ug/L)	0.25

3.0 Calculating the concentration (C) of an element in soil using data from prep log, run log, and quantitation report (note: the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (ug/L)	0.1
Vf = Final volume	200
Vi = Initial volume	0.5
D = Dilution factor as a multiplier (10X = 10)	1
Cx = Concentration of element in (ug/kg)	40

4.0 Adjusting the concentration to dry weight:

$$Cdry = \frac{Cx \times 100}{Px}$$

Where:	**Example:**
Cx = Concentration calculated as received (wet basis)	40
Px = Percent solids of sample (%wt)	80
$Cdry$ = Concentration calculated as dry weight (ug/kg)	50

50 ug/kg = 0.050 mg/kg

Perkin Elmer NexION ICP/MS

STANDARDS KEY

QC Std 1 - ICV
QC Std 2 - ICB
QC Std 3 - LLICV
QC Std 4 - ICSA
QC Std 5 - ICSAB
QC Std 6 - CCV
QC Std 7 - CCB
QC Std 8 - LLCCV

Calibration Solutions

Analyte	Stock Conc. (mg/L)	S1 (mg/L)	S2 (mg/L)	S3 (mg/L)	S4 (mg/L)
Al	10	0	0.00005	0.05	0.1
Sb	10	0	0.00005	0.05	0.1
As	10	0	0.00005	0.05	0.1
Ba	10	0	0.00005	0.05	0.1
Be	10	0	0.00005	0.05	0.1
Ca	1000	0	0.005	5	10
Cd	10	0	0.0005	0.05	0.1
Cr	10	0	0.0005	0.05	0.1
Co	10	0	0.0005	0.05	0.1
Cu	10	0	0.0005	0.05	0.1
Fe	1000	0	0.005	5	10
Pb	10	0	0.00005	0.05	0.1
Mg	1000	0	0.005	5	10
Mn	10	0	0.00005	0.05	0.1
Ni	10	0	0.00005	0.05	0.1
K	1000	0	0.005	5	10
Se	10	0	0.00005	0.05	0.1
Ag	10	0	0.00005	0.05	0.1
Na	1000	0	0.005	5	10
Tl	10	0	0.00005	0.05	0.1
V	10	0	0.00005	0.05	0.1
U	1000	0	0.00005	0.05	0.1
Zn	10	0	0.00005	0.05	0.1

Microbac Laboratories Inc.
Microwave Digestion Log

Workgroup: WG574242
Analyst: VC
Spike Analyst: VC
Run Date: 06/27/2016 08:30
Method: 3015
Balance: BAL016
Instrument: MW-3
Instrument Start: 06/27/2016 08:39

SOP: ME407 Revison 19
Spike Solution: STD75998
Spike Witness: ERP
HNO3 Lot #: COA18905
40 & 50 ML. DIGESTION TU COA18987
MS Filters- fisher-Lot#R RGT35621

	SAMPLE #	Type	Matrix	Initial Amount	Final Volume	Initial Vessel Wt	Final Vessel Wt	Spike Amount	Due Date
1	WG574242-02	BLANK	1	20 mL	50 mL	182.814 g	182.793 g		
2	WG574242-03	LCS	1	20 mL	50 mL	183.56 g	183.549 g	.25 mL	
3	L16060794-01	SAMP	1	20 mL	50 mL	184.361 g	184.346 g		06/24/16
4	L16060794-02	SAMP	1	20 mL	50 mL	183.398 g	183.379 g		06/24/16
5	L16060794-03	SAMP	1	20 mL	50 mL	184.044 g	184.009 g		06/24/16
6	L16060794-04	SAMP	1	20 mL	50 mL	182.433 g	182.424 g		06/24/16
7	L16060822-02	SAMP	1	20 mL	50 mL	182.575 g	182.563 g		06/24/16
8	L16060822-03	SAMP	1	20 mL	50 mL	182.733 g	182.696 g		06/24/16
9	L16060823-02	SAMP	1	20 mL	50 mL	184.331 g	184.334 g		06/24/16
10	L16060823-03	SAMP	1	20 mL	50 mL	181.954 g	181.917 g		06/24/16
11	L16060823-04	SAMP	1	20 mL	50 mL	184.816 g	184.809 g		06/24/16
12	WG574242-01	REF	2	20 mL	50 mL	183.706 g	183.7 g		
13	L16060839-01	SAMP	2	20 mL	50 mL	183.706 g	183.7 g		06/22/16
14	L16060843-01	SAMP	2	20 mL	50 mL	179.215 g	179.196 g		06/22/16
15	L16060843-02	SAMP	2	20 mL	50 mL	181.896 g	181.872 g		06/22/16
16	L16061410-01	SAMP	1	20 mL	50 mL	181.139 g	181.127 g		07/01/16
17	L16061412-01	SAMP	1	20 mL	50 mL	181.093 g	181.089 g		07/01/16
18	L16061412-02	SAMP	1	20 mL	50 mL	184.474 g	184.466 g		07/01/16
19	L16061413-01	SAMP	1	20 mL	50 mL	181.294 g	181.262 g		07/01/16
20	L16061416-01	SAMP	1	20 mL	50 mL	183.477 g	183.459 g		07/01/16
21	L16061416-02	SAMP	1	20 mL	50 mL	179.518 g	178.772 g		07/01/16
22	L16061421-01	SAMP	1	20 mL	50 mL	183.973 g	183.93 g		07/01/16
23	WG574242-04	DUP	1	20 mL	50 mL	183.873 g	183.86 g		
24	WG574242-05	MS	1	20 mL	50 mL	181.12 g	181.111 g	.25 mL	
25	WG574242-06	MSD	1	20 mL	50 mL	183.984 g	183.969 g	.25 mL	

Analyst: Vicki Collins

Reviewer: Erin Potter

MW_DIG - Modified 09/30/2009
PDF ID: 4825468
Report generated: 06/27/2016 09:45

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062816A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 572184,573475,574522,574570,574571,574573

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	NI.062816.130257	Blank	Blank		1		06/28/16 13:02
2	NI.062816.130609	WG574537-01	Calibration Point		1		06/28/16 13:06
3	NI.062816.130921	WG574537-02	Calibration Point		1		06/28/16 13:09
4	NI.062816.131232	WG574537-03	Calibration Point		1		06/28/16 13:12
5	NI.062816.131544	WG574537-04	Calibration Point		1		06/28/16 13:15
6	NI.062816.131857	WG574537-05	Initial Calibration Verification		1		06/28/16 13:18
7	NI.062816.132210	WG574537-06	Initial Calib Blank		1		06/28/16 13:22
8	NI.062816.132523	WG574537-07	Low Level Initial Calibration V		1		06/28/16 13:25
9	NI.062816.132834	WG574537-08	Interference Check		1		06/28/16 13:28
10	NI.062816.133145	WG574537-09	Interference Check		1		06/28/16 13:31
11	NI.062816.133459	WG574537-10	CCV		1		06/28/16 13:34
12	NI.062816.133810	WG574537-11	CCB		1		06/28/16 13:38
13	NI.062816.134122	L16060362-03	6060227-03		10		06/28/16 13:41
14	NI.062816.134434	L16060362-04	6060227-04		1		06/28/16 13:44
15	NI.062816.134903	L16060362-03	6060227-03	20/50	10		06/28/16 13:49
16	NI.062816.135218	L16060362-04	6060227-04	20/50	1		06/28/16 13:52
17	NI.062816.135925	L16060362-05	6060227-05	20/50	1		06/28/16 13:59
18	NI.062816.140236	L16060362-06	6060227-06	20/50	1		06/28/16 14:02
19	NI.062816.140602	L16060362-07	6060227-07	20/50	1		06/28/16 14:06
20	NI.062816.140914	L16060362-08	6060227-08	20/50	1		06/28/16 14:09
21	NI.062816.141225	L16060362-09	6060227-09	20/50	1		06/28/16 14:12
22	NI.062816.141537	L16060362-10	6060227-10	20/50	1		06/28/16 14:15
23	NI.062816.141850	WG574537-12	CCV		1		06/28/16 14:18
24	NI.062816.142201	WG574537-13	CCB		1		06/28/16 14:22
25	NI.062816.142514	L16060362-11	6060227-11	20/50	1		06/28/16 14:25
26	NI.062816.142825	L16060362-12	6060227-12	20/50	1		06/28/16 14:28
27	NI.062816.143138	WG574537-14	CCV		1		06/28/16 14:31
28	NI.062816.143450	WG574537-15	CCB		1		06/28/16 14:34
29	NI.062816.143803	WG573333-03	Method/Prep Blank	.25/100	1		06/28/16 14:38
30	NI.062816.144114	WG573333-04	Laboratory Control S	.25/100	1		06/28/16 14:41
31	NI.062816.144426	WG573333-02	Reference Sample		1	L16061053-01	06/28/16 14:44
32	NI.062816.144737	WG573333-07	Matrix Spike	.253/100	1	L16061053-01	06/28/16 14:47
33	NI.062816.145049	WG573333-08	Matrix Spike Duplica	.252/100	1	L16061053-01	06/28/16 14:50
34	NI.062816.145400	L16061048-01	TR0062	.252/100	5		06/28/16 14:54

Page: 1 Approved: June 29, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062816A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 572184,573475,574522,574570,574571,574573

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	NI.062816.145711	WG573475-03	Post Digestion Spike		5	L16061048-01	06/28/16 14:57
36	NI.062816.150023	WG573475-04	Serial Dilution		25	L16061048-01	06/28/16 15:00
37	NI.062816.150336	WG574537-16	Interference Check		1		06/28/16 15:03
38	NI.062816.150647	WG574537-17	Interference Check		1		06/28/16 15:06
39	NI.062816.151001	WG574537-18	CCV		1		06/28/16 15:10
40	NI.062816.151312	WG574537-19	CCB		1		06/28/16 15:13
41	NI.062816.151629	WG574537-20	Low Level Continuing Calibra		1		06/28/16 15:16
42	NI.062816.152424	WG573333-01	Reference Sample		10	L16060787-03	06/28/16 15:24
43	NI.062816.152735	WG573333-05	Matrix Spike	.256/100	10	L16060787-03	06/28/16 15:27
44	NI.062816.153046	WG573333-06	Matrix Spike Duplica	.253/100	10	L16060787-03	06/28/16 15:30
45	NI.062816.153358	L16060787-06	U517SS00512A	.256/100	10		06/28/16 15:33
46	NI.062816.153709	L16060787-08	U517SS00612A	.254/100	10		06/28/16 15:37
47	NI.062816.154021	L16060787-13	U517SS00708A	.252/100	10		06/28/16 15:40
48	NI.062816.154332	L16060787-15	U517SS00808A	.257/100	10		06/28/16 15:43
49	NI.062816.154644	L16060787-16	U517SS00808C	.255/100	10		06/28/16 15:46
50	NI.062816.154955	L16060787-17	U517SS00810A	.259/100	10		06/28/16 15:49
51	NI.062816.155308	WG574537-21	CCV		1		06/28/16 15:53
52	NI.062816.155620	WG574537-22	CCB		1		06/28/16 15:56
53	NI.062816.155933	WG573751-04	Method/Prep Blank	20/50	1		06/28/16 15:59
54	NI.062816.160244	WG573751-05	Laboratory Control S	20/50	1		06/28/16 16:02
55	NI.062816.160556	WG573751-02	Reference Sample		1	L16061151-01	06/28/16 16:05
56	NI.062816.160907	WG573751-07	Matrix Spike	20/50	1	L16061151-01	06/28/16 16:09
57	NI.062816.161219	WG573751-08	Matrix Spike Duplica	20/50	1	L16061151-01	06/28/16 16:12
58	NI.062816.161530	WG573751-01	Reference Sample		1	L16060916-02	06/28/16 16:15
59	NI.062816.161841	WG573751-06	Duplicate	20/50	1	L16060916-02	06/28/16 16:18
60	NI.062816.162153	L16060916-01	PHL 1 FRESHWATER IMPO	20/50	1		06/28/16 16:21
61	NI.062816.162504	WG574522-01	Post Digestion Spike		1	L16060916-01	06/28/16 16:25
62	NI.062816.162816	WG574522-02	Serial Dilution		5	L16060916-01	06/28/16 16:28
63	NI.062816.163129	WG574537-23	CCV		1		06/28/16 16:31
64	NI.062816.163441	WG574537-24	CCB		1		06/28/16 16:34
65	NI.062816.163754	WG574522-02	Serial Dilution		25	L16060916-01	06/28/16 16:37
66	NI.062816.164105	L16061142-01	18WW20-062016	20/50	1		06/28/16 16:41
67	NI.062816.164417	L16061142-03	18WW19F-062016	20/50	1		06/28/16 16:44
68	NI.062816.164728	L16061142-05	18WW19FFD-062016	20/50	1		06/28/16 16:47

Page: 2 Approved: June 29, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062816A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 572184,573475,574522,574570,574571,574573

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
69	NI.062816.165039	L16061151-02	LF003SW0311-007 DUP	20/50	1		06/28/16 16:50
70	NI.062816.165351	L16061151-05	LF003SW0312-007	20/50	1		06/28/16 16:53
71	NI.062816.165702	WG573751-03	Reference Sample		1	L16061154-01	06/28/16 16:57
72	NI.062816.170014	WG573751-09	Matrix Spike	20/50	1	L16061154-01	06/28/16 17:00
73	NI.062816.170325	WG573751-10	Matrix Spike Duplica	20/50	1	L16061154-01	06/28/16 17:03
74	NI.062816.170637	L16061154-04	FQC-RB06201601-CB	20/50	1		06/28/16 17:06
75	NI.062816.170950	WG574537-25	CCV		1		06/28/16 17:09
76	NI.062816.171301	WG574537-26	CCB		1		06/28/16 17:13
77	NI.062816.171614	L16061155-01	ST008MW0812-007	20/50	1		06/28/16 17:16
78	NI.062816.171926	L16061155-02	ST008MW0809-007	20/50	1		06/28/16 17:19
79	NI.062816.172237	L16061155-03	ST008MW0816-007	20/50	1		06/28/16 17:22
80	NI.062816.172550	WG574537-27	CCV		1		06/28/16 17:25
81	NI.062816.172902	WG574537-28	CCB		1		06/28/16 17:29
82	NI.062816.184632	L16061142-01	18WW20-062016	20/50	10		06/28/16 18:46
83	NI.062816.184946	WG574537-29	CCV		1		06/28/16 18:49
84	NI.062816.185257	WG574537-30	CCB		1		06/28/16 18:52
85	NI.062816.185610	WG574537-31	Low Level Continuing Calibra		1		06/28/16 18:56
86	NI.062816.195129	WG574242-02	Method/Prep Blank	20/50	1		06/28/16 19:51
87	NI.062816.195440	WG574242-03	Laboratory Control S	20/50	1		06/28/16 19:54
88	NI.062816.195752	WG574242-01	Reference Sample		1	L16060839-01	06/28/16 19:57
89	NI.062816.200103	WG574242-04	Duplicate	20/50	1	L16060839-01	06/28/16 20:01
90	NI.062816.200414	WG574242-05	Matrix Spike	20/50	1	L16060839-01	06/28/16 20:04
91	NI.062816.200726	WG574242-06	Matrix Spike Duplica	20/50	1	L16060839-01	06/28/16 20:07
92	NI.062816.201037	L16060843-01	6060966-01	20/50	1		06/28/16 20:10
93	NI.062816.201349	WG574570-01	Post Digestion Spike		1	L16060843-01	06/28/16 20:13
94	NI.062816.201700	WG574570-02	Serial Dilution		5	L16060843-01	06/28/16 20:17
95	NI.062816.202012	WG574570-02	Serial Dilution		25	L16060843-01	06/28/16 20:20
96	NI.062816.202325	WG574537-32	CCV		1		06/28/16 20:23
97	NI.062816.202637	WG574537-33	CCB		1		06/28/16 20:26
98	NI.062816.202950	L16060794-01	A08-MW06-Y1S2	20/50	1		06/28/16 20:29
99	NI.062816.203301	L16060794-02	A08-MW09-Y1S2	20/50	1		06/28/16 20:33
100	NI.062816.203613	L16060794-03	A09-MW02-Y1S2	20/50	1		06/28/16 20:36
101	NI.062816.203924	L16060794-04	A09-MW07-Y1S2	20/50	1		06/28/16 20:39
102	NI.062816.204236	L16060822-02	18WW03F-061416	20/50	1		06/28/16 20:42

Page: 3 Approved: June 29, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062816A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 572184,573475,574522,574570,574571,574573

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
103	NI.062816.204547	L16060822-03	18CPTMW04SW-061416	20/50	1		06/28/16 20:45
104	NI.062816.204859	L16060823-02	126F-061416	20/50	1		06/28/16 20:48
105	NI.062816.205210	L16060823-03	C08-061416	20/50	1		06/28/16 20:52
106	NI.062816.205522	L16060823-04	C03-061416	20/50	1		06/28/16 20:55
107	NI.062816.205833	L16060843-02	6060966-02	20/50	1		06/28/16 20:58
108	NI.062816.210146	WG574537-34	CCV		1		06/28/16 21:01
109	NI.062816.210457	WG574537-35	CCB		1		06/28/16 21:04
110	NI.062816.210810	L16061410-01	SS007MW0704-007	20/50	1		06/28/16 21:08
111	NI.062816.211122	L16061412-01	SS007MW0B05-007	20/50	1		06/28/16 21:11
112	NI.062816.211433	L16061412-02	SS007MW0B02-007	20/50	1		06/28/16 21:14
113	NI.062816.211745	L16061413-01	DP005MW0504-007	20/50	1		06/28/16 21:17
114	NI.062816.212056	L16061416-01	DP005MW0509-007	20/50	1		06/28/16 21:20
115	NI.062816.212408	L16061416-02	DP005MW0518-007	20/50	1		06/28/16 21:24
116	NI.062816.212719	L16061421-01	LF003MW03	20/50	1		06/28/16 21:27
117	NI.062816.213032	WG574537-36	CCV		1		06/28/16 21:30
118	NI.062816.213343	WG574537-37	CCB		1		06/28/16 21:33
119	NI.062816.213655	WG574205-02	Method/Prep Blank	20/50	1		06/28/16 21:36
120	NI.062816.214007	WG574205-03	Laboratory Control S	20/50	1		06/28/16 21:40
121	NI.062816.214318	WG574205-01	Reference Sample		1	L16061343-01	06/28/16 21:43
122	NI.062816.214630	WG574205-04	Duplicate	20/50	1	L16061343-01	06/28/16 21:46
123	NI.062816.214941	WG574205-05	Matrix Spike	20/50	1	L16061343-01	06/28/16 21:49
124	NI.062816.215253	WG574205-06	Matrix Spike Duplica	20/50	1	L16061343-01	06/28/16 21:52
125	NI.062816.215604	L16060732-01	LH18/24-SP650-6371-GRAB	20/50	1		06/28/16 21:56
126	NI.062816.215915	WG574571-01	Post Digestion Spike		1	L16060732-01	06/28/16 21:59
127	NI.062816.220227	WG574571-02	Serial Dilution		5	L16060732-01	06/28/16 22:02
128	NI.062816.220538	WG574571-02	Serial Dilution		25	L16060732-01	06/28/16 22:05
129	NI.062816.220851	WG574537-38	CCV		1		06/28/16 22:08
130	NI.062816.221202	WG574537-39	CCB		1		06/28/16 22:12
131	NI.062816.221514	L16060732-02	LH18/24-SP650-6371-COMP	20/50	1		06/28/16 22:15
132	NI.062816.221825	L16060762-01	MW-24-001	20/50	1		06/28/16 22:18
133	NI.062816.222137	L16061380-01	XU936TW003A	20/50	1		06/28/16 22:21
134	NI.062816.222448	L16061380-02	XU936TW003C	20/50	1		06/28/16 22:24
135	NI.062816.222800	L16061381-06	EB02	20/50	1		06/28/16 22:28
136	NI.062816.223111	L16061393-01	S6F0627-01	20/50	1		06/28/16 22:31

Page: 4 Approved: June 29, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062816A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 572184,573475,574522,574570,574571,574573

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
137	NI.062816.223423	L16061393-02	S6F0627-02	20/50	1		06/28/16 22:34
138	NI.062816.223734	L16061393-03	S6F0627-03	20/50	1		06/28/16 22:37
139	NI.062816.224046	L16061393-04	S6F0627-04	20/50	1		06/28/16 22:40
140	NI.062816.224357	L16061393-05	S6F0627-05	20/50	1		06/28/16 22:43
141	NI.062816.224710	WG574537-40	CCV		1		06/28/16 22:47
142	NI.062816.225022	WG574537-41	CCB		1		06/28/16 22:50
143	NI.062816.225335	L16061393-06	S6F0627-06	20/50	1		06/28/16 22:53
144	NI.062816.225646	L16061393-07	S6F0627-07	20/50	1		06/28/16 22:56
145	NI.062816.225958	L16061401-01	DP005MW0501-007	20/50	1		06/28/16 22:59
146	NI.062816.230309	L16061402-01	SS007MW0715-007	20/50	1		06/28/16 23:03
147	NI.062816.230621	L16061404-01	DP005MW0503-007	20/50	1		06/28/16 23:06
148	NI.062816.230932	L16061408-01	DP005MW0508-007	20/50	1		06/28/16 23:09
149	NI.062816.231244	L16061409-01	ST008MW0818-007	20/50	1		06/28/16 23:12
150	NI.062816.231556	WG574537-42	CCV		1		06/28/16 23:15
151	NI.062816.231908	WG574537-43	CCB		1		06/28/16 23:19
152	NI.062816.232221	WG574537-44	Low Level Continuing Calibra		1		06/28/16 23:22
153	NI.062816.232534	WG574306-02	Method/Prep Blank	.25/100	1		06/28/16 23:25
154	NI.062816.232845	WG574306-03	Laboratory Control S	.25/100	1		06/28/16 23:28
155	NI.062816.233157	L16061297-07	XU936DS01700A		1	WG574306-01	06/28/16 23:31
156	NI.062816.233508	L16061297-08	XU936DS01700MS	.253/100	1	WG574306-04	06/28/16 23:35
157	NI.062816.233819	L16061297-09	XU936DS01700MSD	.255/100	1	WG574306-05	06/28/16 23:38
158	NI.062816.234131	L16061443-01	GS042QACF035SB062416-E	.256/100	1		06/28/16 23:41
159	NI.062816.234442	L16061443-02	GS042QACF035SB062416-E	.256/100	1		06/28/16 23:44
160	NI.062816.234754	WG574573-01	Post Digestion Spike		1	L16061443-02	06/28/16 23:47
161	NI.062816.235105	WG574573-02	Serial Dilution		5	L16061443-02	06/28/16 23:51
162	NI.062816.235417	WG574573-02	Serial Dilution		25	L16061443-02	06/28/16 23:54
163	NI.062816.235730	WG574537-45	CCV		1		06/28/16 23:57
164	NI.062916.000041	WG574537-46	CCB		1		06/29/16 00:00

Page: 5 Approved: June 29, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	NI.062916.112234	Blank	Blank		1		06/29/16 11:22
2	NI.062916.112545	WG574676-01	Calibration Point		1		06/29/16 11:25
3	NI.062916.112857	WG574676-02	Calibration Point		1		06/29/16 11:28
4	NI.062916.113208	WG574676-03	Calibration Point		1		06/29/16 11:32
5	NI.062916.113520	WG574676-04	Calibration Point		1		06/29/16 11:35
6	NI.062916.113833	WG574676-05	Initial Calibration Verification		1		06/29/16 11:38
7	NI.062916.114146	WG574676-06	Initial Calib Blank		1		06/29/16 11:41
8	NI.062916.114653	WG574676-07	Low Level Initial Calibration V		1		06/29/16 11:46
9	NI.062916.115053	WG574676-08	Interference Check		1		06/29/16 11:50
10	NI.062916.115447	WG574676-09	Interference Check		1		06/29/16 11:54
11	NI.062916.120010	WG574676-10	CCV		1		06/29/16 12:00
12	NI.062916.120734	WG574676-11	CCB		1		06/29/16 12:07
13	NI.062916.121215	L16060661-06	US23SS02502A	.253/100	1		06/29/16 12:12
14	NI.062916.121526	L16060661-07	US23SS02505A	.257/100	1		06/29/16 12:15
15	NI.062916.121838	L16060661-08	US23SS02602A	.257/100	1		06/29/16 12:18
16	NI.062916.122149	L16060661-09	US23SS02605A	.25/100	1		06/29/16 12:21
17	NI.062916.122501	L16060822-03	18CPTMW04SW-061416	20/50	10		06/29/16 12:25
18	NI.062916.122812	L16060823-02	126F-061416	20/50	50		06/29/16 12:28
19	NI.062916.123124	L16060823-03	C08-061416	20/50	50		06/29/16 12:31
20	NI.062916.123435	L16060823-04	C03-061416	20/50	50		06/29/16 12:34
21	NI.062916.123748	WG574676-12	CCV		1		06/29/16 12:37
22	NI.062916.124059	WG574676-13	CCB		1		06/29/16 12:40
23	NI.062916.124412	WG574676-14	Low Level Continuing Calibra		1		06/29/16 12:44
24	NI.062916.125354	WG573548-02	Method/Prep Blank	20/50	1		06/29/16 12:53
25	NI.062916.125705	WG573548-03	Laboratory Control S	20/50	1		06/29/16 12:57
26	NI.062916.130017	WG573548-01	Reference Sample		1	L16060986-23	06/29/16 13:00
27	NI.062916.130328	WG573548-04	Matrix Spike	20/50	1	L16060986-23	06/29/16 13:03
28	NI.062916.130639	WG573548-05	Matrix Spike Duplica	20/50	1	L16060986-23	06/29/16 13:06
29	NI.062916.130951	L16060986-01	MW-07-04	20/50	1		06/29/16 13:09
30	NI.062916.131302	L16060986-03	MW-01-10	20/50	1		06/29/16 13:13
31	NI.062916.131614	WG574572-01	Post Digestion Spike		1	L16060986-03	06/29/16 13:16
32	NI.062916.131925	WG574572-02	Serial Dilution		5	L16060986-03	06/29/16 13:19
33	NI.062916.132237	WG574572-02	Serial Dilution		25	L16060986-03	06/29/16 13:22
34	NI.062916.132550	WG574676-15	CCV		1		06/29/16 13:25

Page: 1 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	NI.062916.132901	WG574676-16	CCB		1		06/29/16 13:29
36	NI.062916.133214	L16060986-05	MW-03-12	20/50	1		06/29/16 13:32
37	NI.062916.133526	L16060986-07	MW-01-07	20/50	1		06/29/16 13:35
38	NI.062916.133837	L16060986-09	MW-03-11	20/50	1		06/29/16 13:38
39	NI.062916.134150	L16060986-11	MW-01-01	20/50	1		06/29/16 13:41
40	NI.062916.134502	L16060986-13	MW-03-09	20/50	1		06/29/16 13:45
41	NI.062916.134813	L16060986-15	MW-01-08	20/50	1		06/29/16 13:48
42	NI.062916.135125	L16060986-17	MW-03-06	20/50	1		06/29/16 13:51
43	NI.062916.135436	L16060986-19	MW-03-05	20/50	1		06/29/16 13:54
44	NI.062916.135748	L16060986-25	MW-01-02	20/50	1		06/29/16 13:57
45	NI.062916.140059	L16060986-27	DUP 1	20/50	1		06/29/16 14:00
46	NI.062916.140412	WG574676-17	CCV		1		06/29/16 14:04
47	NI.062916.140723	WG574676-18	CCB		1		06/29/16 14:07
48	NI.062916.141036	L16060986-29	DUP 2	20/50	1		06/29/16 14:10
49	NI.062916.141347	L16060986-39	EQUIP BLANK 1	20/50	1		06/29/16 14:13
50	NI.062916.141659	L16060986-41	EQUIP BLANK 2	20/50	1		06/29/16 14:16
51	NI.062916.145120	L16061026-02	18WW16F-061616	20/50	1		06/29/16 14:51
52	NI.062916.145431	L16061026-04	129-061616	20/50	1		06/29/16 14:54
53	NI.062916.145742	L16061026-06	120F-061616	20/50	1		06/29/16 14:57
54	NI.062916.150054	L16061026-07	123-061616	20/50	1		06/29/16 15:00
55	NI.062916.150407	WG574676-19	CCV		1		06/29/16 15:04
56	NI.062916.150718	WG574676-20	CCB		1		06/29/16 15:07
57	NI.062916.151032	WG573099-02	Method/Prep Blank	.25/100	1		06/29/16 15:10
58	NI.062916.151343	WG573099-03	Laboratory Control S	.25/100	1		06/29/16 15:13
59	NI.062916.151655	WG573099-01	Reference Sample		10	L16060888-03	06/29/16 15:16
60	NI.062916.152006	WG573099-04	Matrix Spike	.255/100	10	L16060888-03	06/29/16 15:20
61	NI.062916.152318	WG573099-05	Matrix Spike Duplica	.254/100	10	L16060888-03	06/29/16 15:23
62	NI.062916.152629	L16060883-01	WC01	.252/100	1		06/29/16 15:26
63	NI.062916.152941	L16060912-01	WH8011	.25/100	1		06/29/16 15:29
64	NI.062916.153252	WG573717-03	Post Digestion Spike		1	L16060912-01	06/29/16 15:32
65	NI.062916.153604	WG573717-04	Serial Dilution		5	L16060912-01	06/29/16 15:36
66	NI.062916.153915	WG573717-04	Serial Dilution		25	L16060912-01	06/29/16 15:39
67	NI.062916.154228	WG574676-21	CCV		1		06/29/16 15:42
68	NI.062916.154540	WG574676-22	CCB		1		06/29/16 15:45

Page: 2 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
69	NI.062916.154853	WG574001-04	Method/Prep Blank	20/50	1		06/29/16 15:48
70	NI.062916.155204	WG574001-05	Laboratory Control S	20/50	1		06/29/16 15:52
71	NI.062916.155516	WG574001-01	Reference Sample		1	L16061247-02	06/29/16 15:55
72	NI.062916.155827	WG574001-06	Duplicate	20/50	1	L16061247-02	06/29/16 15:58
73	NI.062916.160139	WG574001-02	Reference Sample		1	L16061265-01	06/29/16 16:01
74	NI.062916.160450	WG574001-07	Matrix Spike	20/50	1	L16061265-01	06/29/16 16:04
75	NI.062916.160802	WG574001-08	Matrix Spike Duplica	20/50	1	L16061265-01	06/29/16 16:08
76	NI.062916.161113	L16061247-01	WMB09	20/50	1		06/29/16 16:11
77	NI.062916.161425	WG574688-01	Post Digestion Spike		1	L16061247-01	06/29/16 16:14
78	NI.062916.161736	WG574688-02	Serial Dilution		5	L16061247-01	06/29/16 16:17
79	NI.062916.162049	WG574676-23	CCV		1		06/29/16 16:20
80	NI.062916.162401	WG574676-24	CCB		1		06/29/16 16:24
81	NI.062916.162713	WG574688-02	Serial Dilution		25	L16061247-01	06/29/16 16:27
82	NI.062916.163025	L16061265-04	SS007MWOB16-007	20/50	1		06/29/16 16:30
83	NI.062916.163336	L16061265-05	SS007MWOB21-007	20/50	1		06/29/16 16:33
84	NI.062916.163648	L16061265-06	FQC-RB06211601-CC	20/50	1		06/29/16 16:36
85	NI.062916.163959	L16061266-01	ST008MW0805-007	20/50	1		06/29/16 16:39
86	NI.062916.164310	L16061269-01	FT002MW0204-007	20/50	1		06/29/16 16:43
87	NI.062916.164622	L16061272-01	SS007MW0B18-007	20/50	1		06/29/16 16:46
88	NI.062916.164933	L16061272-02	SS007MWN752-007	20/50	1		06/29/16 16:49
89	NI.062916.165245	L16061279-01	XU936TW007A	20/50	1		06/29/16 16:52
90	NI.062916.165556	L16061279-01	XU936TW007A	20/50	50		06/29/16 16:55
91	NI.062916.165909	WG574676-25	CCV		1		06/29/16 16:59
92	NI.062916.170220	WG574676-26	CCB		1		06/29/16 17:02
93	NI.062916.170533	WG574001-03	Reference Sample		1	L16061279-02	06/29/16 17:05
94	NI.062916.170845	WG574001-03	Reference Sample		50	L16061279-02	06/29/16 17:08
95	NI.062916.171156	WG574001-09	Matrix Spike	20/50	1	L16061279-02	06/29/16 17:11
96	NI.062916.171508	WG574001-09	Matrix Spike	20/50	50	L16061279-02	06/29/16 17:15
97	NI.062916.171819	WG574001-10	Matrix Spike Duplica	20/50	1	L16061279-02	06/29/16 17:18
98	NI.062916.172131	WG574001-10	Matrix Spike Duplica	20/50	50	L16061279-02	06/29/16 17:21
99	NI.062916.172442	L16061279-05	XU936TW019A	20/50	1		06/29/16 17:24
100	NI.062916.172754	L16061279-05	XU936TW019A	20/50	50		06/29/16 17:27
101	NI.062916.173105	L16061281-01	J6F0803-01	20/50	1		06/29/16 17:31
102	NI.062916.173419	WG574676-27	CCV		1		06/29/16 17:34

Page: 3 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
103	NI.062916.173730	WG574676-28	CCB		1		06/29/16 17:37
104	NI.062916.174043	L16061294-03	EB01	20/50	1		06/29/16 17:40
105	NI.062916.174355	L16061294-03	EB01	20/50	50		06/29/16 17:43
106	NI.062916.174706	L16061296-01	OUTFALL 002/COMP	20/50	1		06/29/16 17:47
107	NI.062916.175019	WG574676-29	Interference Check		1		06/29/16 17:50
108	NI.062916.175331	WG574676-30	Interference Check		1		06/29/16 17:53
109	NI.062916.175644	WG574676-31	CCV		1		06/29/16 17:56
110	NI.062916.175955	WG574676-32	CCB		1		06/29/16 17:59
111	NI.062916.180308	WG574676-33	Low Level Continuing Calibra		1		06/29/16 18:03
112	NI.062916.180621	WG574192-02	Method/Prep Blank	.25/100	1		06/29/16 18:06
113	NI.062916.180932	WG574192-03	Laboratory Control S	.25/100	1		06/29/16 18:09
114	NI.062916.181244	WG574192-01	Reference Sample		1	L16061206-01	06/29/16 18:12
115	NI.062916.181555	L16061206-02	LF003SD0311-007 DUP	.251/100	1		06/29/16 18:15
116	NI.062916.181907	WG574192-04	Matrix Spike	.25/100	1	L16061206-01	06/29/16 18:19
117	NI.062916.182218	WG574192-05	Matrix Spike Duplica	.253/100	1	L16061206-01	06/29/16 18:22
118	NI.062916.182530	L16061206-05	LF003SD0312-007	.25/100	1		06/29/16 18:25
119	NI.062916.182841	WG574690-01	Post Digestion Spike		1	L16061206-05	06/29/16 18:28
120	NI.062916.183153	WG574690-02	Serial Dilution		5	L16061206-05	06/29/16 18:31
121	NI.062916.183504	WG574690-02	Serial Dilution		25	L16061206-05	06/29/16 18:35
122	NI.062916.183818	WG574676-34	CCV		1		06/29/16 18:38
123	NI.062916.184129	WG574676-35	CCB		1		06/29/16 18:41
124	NI.062916.184442	L16061207-01	SS007SD0709-007	.251/100	1		06/29/16 18:44
125	NI.062916.184754	L16061207-02	SS007SD0710-007	.25/100	1		06/29/16 18:47
126	NI.062916.185105	L16061208-01	DP005SD0513-007	.252/100	1		06/29/16 18:51
127	NI.062916.185417	L16061208-02	DP005SD0514-007	.251/100	1		06/29/16 18:54
128	NI.062916.185728	L16061291-01	V6F0270-01	.253/100	1		06/29/16 18:57
129	NI.062916.190040	L16061381-01	XU936DS00500A	.254/100	1		06/29/16 19:00
130	NI.062916.190351	L16061381-01	XU936DS00500A	.254/100	50		06/29/16 19:03
131	NI.062916.190703	L16061381-02	XU936DS00505A	.258/100	1		06/29/16 19:07
132	NI.062916.191014	L16061381-02	XU936DS00505A	.258/100	50		06/29/16 19:10
133	NI.062916.191327	WG574676-36	CCV		1		06/29/16 19:13
134	NI.062916.191639	WG574676-37	CCB		1		06/29/16 19:16
135	NI.062916.191952	L16061381-03	XU936DS00300A	.252/100	1		06/29/16 19:19
136	NI.062916.192303	L16061381-03	XU936DS00300A	.252/100	50		06/29/16 19:23

Page: 4 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP

Analyst1: BKT Analyst2: N/A

Method: 6020/6020A/200.8 SOP: ME700A Rev: 2

Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705

ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607

CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345

Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
137	NI.062916.192615	L16061381-04	XU936DS00305A	.25/100	1		06/29/16 19:26
138	NI.062916.192926	L16061381-04	XU936DS00305A	.25/100	50		06/29/16 19:29
139	NI.062916.193237	L16061381-05	XU936DS00305C	.253/100	1		06/29/16 19:32
140	NI.062916.193550	L16061381-05	XU936DS00305C	.253/100	50		06/29/16 19:35
141	NI.062916.193901	L16061381-07	XU936DS00400A	.252/100	1		06/29/16 19:39
142	NI.062916.194213	L16061381-07	XU936DS00400A	.252/100	50		06/29/16 19:42
143	NI.062916.194524	L16061384-02	BFB-16-019	.253/100	1		06/29/16 19:45
144	NI.062916.194835	L16061384-04	BFB-16-020	.253/100	1		06/29/16 19:48
145	NI.062916.195149	WG574676-38	CCV		1		06/29/16 19:51
146	NI.062916.195500	WG574676-39	CCB		1		06/29/16 19:55
147	NI.062916.195813	L16061384-06	BFB-16-021	.253/100	1		06/29/16 19:58
148	NI.062916.200124	L16061384-08	BFB-16-022	.251/100	1		06/29/16 20:01
149	NI.062916.200436	L16061384-10	BFB-16-023	.254/100	1		06/29/16 20:04
150	NI.062916.200747	L16061384-12	BFB-16-024	.254/100	1		06/29/16 20:07
151	NI.062916.201101	WG574676-40	CCV		1		06/29/16 20:11
152	NI.062916.201412	WG574676-41	CCB		1		06/29/16 20:14
153	NI.062916.201726	WG573291-03	Method/Prep Blank	20/50	1		06/29/16 20:17
154	NI.062916.202037	WG573291-04	Laboratory Control S	20/50	1		06/29/16 20:20
155	NI.062916.202348	WG573291-01	Reference Sample		1	L16060974-01	06/29/16 20:23
156	NI.062916.202700	WG573291-05	Matrix Spike	20/50	1	L16060974-01	06/29/16 20:27
157	NI.062916.203011	WG573291-06	Matrix Spike Duplica	20/50	1	L16060974-01	06/29/16 20:30
158	NI.062916.203323	WG573790-03	Post Digestion Spike		1	L16060974-01	06/29/16 20:33
159	NI.062916.203634	WG573790-04	Serial Dilution		5	L16060974-01	06/29/16 20:36
160	NI.062916.203946	WG573790-04	Serial Dilution		25	L16060974-01	06/29/16 20:39
161	NI.062916.204259	WG574676-42	CCV		1		06/29/16 20:42
162	NI.062916.204611	WG574676-43	CCB		1		06/29/16 20:46

Page: 5 Approved: July 05, 2016

Microbac Laboratories Inc.

Data Checklist

Date: 28-JUN-2016
Analyst: BKT
Analyst: NA
Method: 6020/6020A/200.8
Instrument: ICP-MS2
Curve Workgroup: 574537
Runlog ID: 75986
Analytical Workgroups: 572184,573475,574522,574570,574571,574573

STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	
ICB/CCB	X
ICSA/ICSAB	X
CRI	
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	X
Client Forms	X
Level X	
Level 3	762,1380,1381
Level 4	732,787,822,823,1048,1142,1151,1154
Level 4	1155,1401,1402,1404,1408,1409,1410
Level 4	1412,1413,1416,1421,1443
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	BKT
Secondary Reviewer	KHR
Comments	

Primary Reviewer:
29-JUN-2016

Secondary Reviewer:
29-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-29-2016 15:32:08

Microbac Laboratories Inc.

Data Checklist

Date: 29-JUN-2016
Analyst: BKT
Analyst: NA
Method: 6020/6020A/200.8
Instrument: ICP-MS2
Curve Workgroup: 574676
Runlog ID: 76013
Analytical Workgroups: 573722,574570,574572,573717,574688,574690,573790

STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	
ICB/CCB	X
ICSA/ICSAB	X
CRI	
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	X
Client Forms	X
Level X	
Level 3	1279,1294,1381,1384
Level 4	661,822,823,883,912,974,986,1026
Level 4	1206,1207,1208,1265,1266,1269,1272
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	BKT
Secondary Reviewer	JYH
Comments	

Primary Reviewer:
04-JUL-2016

Secondary Reviewer:
05-JUL-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUL-05-2016 12:36:33

Microbac Laboratories Inc.

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 6020A

AAB#: WG574570

Login Number: L16060822

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18WW03F-061416	02	06/14/16					06/27/2016	13	180		06/28/16	14.5	180	
18CPTMW04SW-061416	03	06/14/16					06/27/2016	13	180		06/29/16	15.1	180	
18CPTMW04SW-061416	03	06/14/16					06/27/2016	13	180		06/28/16	14.5	180	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4829693
Report generated 06/30/2016 09:23

METHOD BLANK SUMMARY

Login Number: L16060822
Blank File ID: NI.062816.195129
Prep Date: 06/27/16 08:30
Analyzed Date: 06/28/16 19:51
Analyst: BKT

Work Group: WG574570
Blank Sample ID: WG574242-02
Instrument ID: ICP-MS2
Method: 6020A

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG574242-03	NI.062816.195440	06/28/16 19:54	01
DUP	WG574242-04	NI.062816.200103	06/28/16 20:01	01
18WW03F-061416	L16060822-02	NI.062816.204236	06/28/16 20:42	02
18CPTMW04SW-061416	L16060822-03	NI.062816.204547	06/28/16 20:45	02
18CPTMW04SW-061416	L16060822-03	NI.062916.122501	06/29/16 12:25	DL01

Report Name: BLANK_SUMMARY
PDF File ID: 4829694
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060822 Prep Date: 06/27/16 08:30 Sample ID: WG574242-02
Instrument ID: ICP-MS2 Run Date: 06/28/16 19:51 Prep Method: 3015
File ID: NI.062816.195129 Analyst: BKT Method: 6020A
Workgroup (AAB#): WG574570 Matrix: Water Units: mg/L
Contract #: Cal ID: ICP-MS-28-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Antimony, Total	0.000500	0.00200	0.000500	1	U
Arsenic, Total	0.000500	0.00200	0.000500	1	U
Barium, Total	0.00150	0.00600	0.00150	1	U
Cadmium, Total	0.000300	0.00120	0.000300	1	U
Chromium, Total	0.00100	0.00400	0.00100	1	U
Cobalt, Total	0.000500	0.00200	0.000500	1	U
Copper, Total	0.00100	0.00400	0.00100	1	U
Lead, Total	0.000500	0.00200	0.000500	1	U
Manganese, Total	0.00100	0.00400	0.00100	1	U
Nickel, Total	0.00200	0.00800	0.00200	1	U
Silver, Total	0.000500	0.00200	0.000500	1	U
Thallium, Total	0.000100	0.000400	0.000100	1	U
Vanadium, Total	0.000500	0.00200	0.000500	1	U
Zinc, Total	0.0125	0.0500	0.0125	1	U

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4829695
29-JUN-2016 08:34

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060822 Run Date: 06/28/2016 Sample ID: WG574242-03
Instrument ID: ICP-MS2 Run Time: 19:54 Prep Method: 3015
File ID: NI.062816.195440 Analyst: BKT Method: 6020A
Workgroup (AAB#): WG574570 Matrix: Water Units: mg/L
QC Key: DOD4 Lot#: STD75998 Cal ID: ICP-MS-28-JUN-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Antimony, Total	0.125	0.120	96.4	80 - 120	
Arsenic, Total	0.125	0.124	99.0	80 - 120	
Barium, Total	0.125	0.121	96.5	80 - 120	
Cadmium, Total	0.125	0.120	95.7	80 - 120	
Chromium, Total	0.125	0.125	100	80 - 120	
Cobalt, Total	0.125	0.122	97.6	80 - 120	
Copper, Total	0.125	0.127	102	80 - 120	
Lead, Total	0.125	0.125	99.8	80 - 120	
Manganese, Total	0.125	0.128	103	80 - 120	
Nickel, Total	0.125	0.126	100	80 - 120	
Silver, Total	0.125	0.123	98.2	80 - 120	
Thallium, Total	0.125	0.122	97.3	80 - 120	
Vanadium, Total	0.125	0.123	98.1	80 - 120	
Zinc, Total	0.125	0.129	103	80 - 120	

LCS - Modified 03/06/2008
PDF File ID: 4829696
Report generated: 06/29/2016 08:34

Microbac Laboratories Inc.
MATRIX SPIKE AND MATRIX SPIKE DUP (MS/MSD)

Loginnum: L16060822 | Cal ID: ICP-MS2- | Worknum: WG574570
Instrument ID: ICP-MS2 | Contract #: | Method: 6020A
Parent ID: WG574242-01 | File ID: NI.062816.195752 | Dil: 1 | Matrix: WATER
Sample ID: WG574242-05 MS | File ID: NI.062816.200414 | Dil: 1 | Units: mg/L
Sample ID: WG574242-06 MSD | File ID: NI.062816.200726 | Dil: 1

Analyte	Parent	MS Spiked	MS Found	MS %Rec	MSD Spiked	MSD Found	MSD %Rec	%RPD	%Rec Limits	RPD Limit	Q
Antimony	0.000998	0.125	0.123	97.9	0.125	0.120	95.5	2.49	80 - 120	20	
Arsenic	0.00141	0.125	0.127	101	0.125	0.126	99.9	0.561	80 - 120	20	
Barium	0.0414	0.125	0.160	95.3	0.125	0.156	92.0	2.54	80 - 120	20	
Cadmium	ND	0.125	0.121	97.2	0.125	0.118	94.0	3.31	80 - 120	20	
Chromium	ND	0.125	0.121	96.6	0.125	0.124	99.2	2.65	80 - 120	20	
Cobalt	ND	0.125	0.121	97.2	0.125	0.120	96.3	0.931	80 - 120	20	
Copper	0.00439	0.125	0.124	95.7	0.125	0.125	96.4	0.655	80 - 120	20	
Lead	0.00115	0.125	0.126	100	0.125	0.124	98.6	1.64	80 - 120	20	
Manganese	0.00611	0.125	0.129	98.4	0.125	0.129	98.3	0.106	80 - 120	20	
Nickel	0.00410	0.125	0.123	95.0	0.125	0.125	96.9	1.96	80 - 120	20	
Silver	ND	0.125	0.122	97.6	0.125	0.118	94.3	3.46	80 - 120	20	
Thallium	ND	0.125	0.123	98.5	0.125	0.121	97.0	1.54	80 - 120	20	
Vanadium	0.00107	0.125	0.124	98.5	0.125	0.123	97.8	0.621	80 - 120	20	
Zinc	0.0337	0.125	0.154	96.2	0.125	0.156	97.7	1.20	80 - 120	20	

* FAILS %REC LIMIT

FAILS RPD LIMIT

NOTE: This is an internal quality control sample.

WG_MS_MSD_DRYWT - Modified 05/26/2011
PDF File ID: 4829697
Report generated 06/29/2016 08:34

Microbac Laboratories Inc.
Serial Dilution Report

Login: L16060822 **Worknum:** WG574570
Instrument: ICP-MS2 **Method:** 6020A
Serial Dil: WG574570-02 **File ID:** NI.062816.201700 **Dil:** 5 **Units:** ug/L
Sample: L16060843-01 **File ID:** NI.062816.201037 **Dil:** 1

Analyte	Sample	Qual	Serial Dil	Qual	% Diff	Q
Antimony	0.382	F	2.24	F	486.00	
Arsenic	ND	U	ND	U		
Barium	29.9	X	30.7	X	2.82	
Cadmium	0.151	F	ND	U		
Chromium	ND	U	ND	U		
Cobalt	7.40	X	7.21	X	2.48	
Copper	0.509	F	ND	U		
Lead	ND	U	ND	U		
Manganese	716		689		3.69	
Nickel	17.5	X	17.0	X	2.69	
Silver	ND	U	ND	U		
Thallium	ND	U	ND	U		
Vanadium	ND	U	ND	U		
Zinc	55.5	X	55.3	F	0.51	

U = Result is below MDL.
F = Result is greater than or equal to MDL and less than the RL.
X = Result is greater than or equal to RL and less than 100 times the MDL.
E = %D exceeds control limit of 10% and initial sample result is greater than or equal to 100 times the MDL.

SERIAL_DIL - Modified 09/22/2008
PDF File ID: 4829691
06/29/2016 08:34

Microbac Laboratories Inc.
POST SPIKE REPORT

Sample Login ID: L16060822 **Worknum:** WG574570
Instrument ID: ICP-MS2 **Method:** 6020A
Post Spike ID: WG574570-01 **File ID:** NI.062816.201349 **Dil:** 1 **Units:** ug/L
Sample ID: L16060843-01 **File ID:** NI.062816.201037 **Dil:** 1 **Matrix:** Water

Analyte	Post Spike Result	C	Sample Result	C	Spike Added(SA)	% R	Control Limit %R	Q
ANTIMONY	49.7		0.382	F	50	98.7	75 - 125	
ARSENIC	51.2		0	U	50	102.5	75 - 125	
BARIUM	78.1		29.9		50	96.4	75 - 125	
CADMIUM	49.2		0.151	F	50	98.1	75 - 125	
CHROMIUM	50.1		0	U	50	100.1	75 - 125	
COBALT	59.3		7.40		50	103.7	75 - 125	
COPPER	50.9		0.509	F	50	100.8	75 - 125	
LEAD	52.4		0	U	50	104.7	75 - 125	
MANGANESE	748		716		50	63.6	75 - 125	N
NICKEL	66.6		17.5		50	98.2	75 - 125	
SILVER	48.6		0	U	50	97.3	75 - 125	
THALLIUM	50.9		0	U	50	101.8	75 - 125	
VANADIUM	50.8		0	U	50	101.7	75 - 125	
ZINC	115		55.5		50	119.5	75 - 125	

N = % Recovery exceeds control limits
F = Result is between MDL and RL
U = Sample result is below MDL. A value of zero is used in the calculation

POST_SPIKE - Modified 03/06/2008
PDF File ID: 4829692
Report generated: 06/29/2016 08:34

Microbac Laboratories Inc.
Initial Calibration Summary

Login:	L16060822	Workgroup (AAB#):	WG574570
Analytical Method:	6020A	Instrument ID:	ICP-MS2
ICAL Worknum:	WG574537	Initial Calibration Date:	28-JUN-2016 13:15

	WG574537-01		WG574537-02		WG574537-03		WG574537-04			
	Conc	INT	Conc	INT	Conc	INT	Conc	INT	R	Q
ANTIMONY	0	47.7	.4	303	50	245000	100	482000	.999953	
ARSENIC	0	-33.8	.4	11.1	50	59300	100	117000	.999972	
BARIUM	0	33.7	.4	159	50	107000	100	212000	.999991	
CADMIUM	0	3.30	.4	116	50	110000	100	219000	.999985	
CHROMIUM	0	7410	.4	8070	50	376000	100	727000	.999989	
COBALT	0	227	.4	599	50	370000	100	731000	.999947	
COPPER	0	135	.4	280	50	102000	100	195000	.999984	
LEAD	0	933	.4	2130	50	978000	100	1950000	.999942	
MANGANESE	0	835	.4	1350	50	340000	100	679000	.99986	
NICKEL	0	183	.4	306	50	100000	100	194000	1	
SILVER	0	79.0	.4	455	50	352000	100	679000	.999803	
THALLIUM	0	5.30	.4	473	50	438000	100	866000	.999971	
VANADIUM	0	1490	.4	1870	50	370000	100	719000	.999996	
ZINC	0	520	.4	1010	50	59300	100	114000	.999999	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
*** = Out of Compliance; R < 0.995**

INT_CAL_ICP - Modified 03/06/2008
PDF File ID: 4829700
Report generated: 30-JUN-2016 09:23

Microbac Laboratories Inc.
Initial Calibration Summary

Login:	L16060822	Workgroup (AAB#):	WG574570
Analytical Method:	6020A	Instrument ID:	ICP-MS2
ICAL Worknum:	WG574676	Initial Calibration Date:	29-JUN-2016 11:35

	WG574676-01		WG574676-02		WG574676-03		WG574676-04			
	Conc	INT	Conc	INT	Conc	INT	Conc	INT	R	Q
ANTIMONY	0	28.2	.4	255	50	227000	100	448000	.999981	
ARSENIC	0	-40.0	.4	-10.4	50	57000	100	110000	.999904	
BARIUM	0	25.3	.4	136	50	99600	100	196000	.999974	
CADMIUM	0	1.30	.4	122	50	102000	100	200000	.999946	
CHROMIUM	0	7540	.4	7930	50	379000	100	748000	.999997	
COBALT	0	239	.4	575	50	337000	100	648000	.99988	
COPPER	0	103	.4	285	50	94300	100	184000	.999979	
LEAD	0	873	.4	2280	50	926000	100	1830000	.999949	
MANGANESE	0	784	.4	1380	50	355000	100	686000	.999925	
NICKEL	0	221	.4	324	50	94600	100	186000	.999991	
SILVER	0	86.3	.4	412	50	318000	100	624000	.999961	
THALLIUM	0	1.70	.4	417	50	406000	100	804000	.999963	
VANADIUM	0	1430	.4	1850	50	385000	100	745000	.999935	
ZINC	0	365	.4	549	50	56000	100	108000	.999932	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
*** = Out of Compliance; R < 0.995**

INT_CAL_ICP - Modified 03/06/2008
PDF File ID: 4829700
Report generated: 30-JUN-2016 09:23

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-06
Instrument ID: ICP-MS2
Run Time: 13:22
Method: 6020A
File ID: NI.062816.132210
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS2 - 28-JUN-16
Matrix: WATER

Analytes	MDL	RDL	Concentration	Qualifier
SILVER	.2	.8	.2	U
ARSENIC	.2	.8	.2	U
BARIUM	.6	2.4	.6	U
CADMIUM	.12	.48	.12	U
COBALT	.2	.8	.2	U
CHROMIUM	.4	1.6	.4	U
COPPER	.4	1.6	.4	U
MANGANESE	.4	1.6	.4	U
NICKEL	.8	3.2	.8	U
LEAD	.2	.8	.2	U
ANTIMONY	.2	.8	.405	*
THALLIUM	.04	.16	.04	U
VANADIUM	.2	.8	.2	U
ZINC	5	20	5	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4829702
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16060822
Instrument ID: ICP-MS2
File ID: NI.062916.114146
Workgroup (AAB#): WG574570
Matrix: WATER

Run Date: 06/29/2016
Run Time: 11:41
Analyst: BKT
Cal ID: ICP-MS2 - 29-JUN-16

Sample ID: WG574676-06
Method: 6020A
Units: ug/L

Analytes	MDL	RDL	Concentration	Qualifier
SILVER	.2	.8	.2	U
ARSENIC	.2	.8	.2	U
BARIUM	.6	2.4	.6	U
CADMIUM	.12	.48	.12	U
COBALT	.2	.8	.2	U
CHROMIUM	.4	1.6	.4	U
COPPER	.4	1.6	.4	U
MANGANESE	.4	1.6	.4	U
NICKEL	.8	3.2	.8	U
LEAD	.2	.8	.2	U
ANTIMONY	.2	.8	.335	F
THALLIUM	.04	.16	.04	U
VANADIUM	.2	.8	.2	U
ZINC	5	20	5	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4829702
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-11
Instrument ID: ICP-MS2
Run Time: 13:38
Method: 6020A
File ID: NI.062816.133810
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.259	F
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829705
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-15
Instrument ID: ICP-MS2
Run Time: 14:34
Method: 6020A
File ID: NI.062816.143450
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829705
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-19
Instrument ID: ICP-MS2
Run Time: 15:13
Method: 6020A
File ID: NI.062816.151312
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.263	F
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829705
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-30
Instrument ID: ICP-MS2
Run Time: 18:52
Method: 6020A
File ID: NI.062816.185257
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829705
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-33
Instrument ID: ICP-MS2
Run Time: 20:26
Method: 6020A
File ID: NI.062816.202637
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829705
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-35
Instrument ID: ICP-MS2
Run Time: 21:04
Method: 6020A
File ID: NI.062816.210457
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829705
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/29/2016
Sample ID: WG574676-11
Instrument ID: ICP-MS2
Run Time: 12:07
Method: 6020A
File ID: NI.062916.120734
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829705
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/29/2016
Sample ID: WG574676-13
Instrument ID: ICP-MS2
Run Time: 12:40
Method: 6020A
File ID: NI.062916.124059
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829705
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.

INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16060822
Run Date: 06/29/2016
Sample ID: WG574676-05
Instrument ID: ICP-MS2
Run Time: 11:38
Method: 6020A
File ID: NI.062916.113833
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Antimony		50	49.8	99.7	90 - 110	
Arsenic		50	50.9	102	90 - 110	
Barium		50	49.9	99.9	90 - 110	
Cadmium		50	50.3	101	90 - 110	
Chromium		50	50.0	100	90 - 110	
Cobalt		50	50.4	101	90 - 110	
Copper		50	50.8	102	90 - 110	
Lead		50	51.0	102	90 - 110	
Manganese		50	51.9	104	90 - 110	
Nickel		50	50.6	101	90 - 110	
Silver		50	50.9	102	90 - 110	
Thallium		50	50.7	101	90 - 110	
Vanadium		50	51.2	102	90 - 110	
Zinc		50	52.1	104	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4829701
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.

INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16060822 Run Date: 06/28/2016 Sample ID: WG574537-05
Instrument ID: ICP-MS2 Run Time: 13:18 Method: 6020A
File ID: NI.062816.131857 Analyst: BKT Units: ug/L
Workgroup (AAB#): WG574570 Cal ID: ICP-MS - 28-JUN-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Antimony		50	48.3	96.6	90 - 110	
Arsenic		50	50.3	101	90 - 110	
Barium		50	48.2	96.4	90 - 110	
Cadmium		50	49.0	98.0	90 - 110	
Chromium		50	50.0	100	90 - 110	
Cobalt		50	50.2	100	90 - 110	
Copper		50	50.1	100	90 - 110	
Lead		50	48.9	97.7	90 - 110	
Manganese		50	50.2	100	90 - 110	
Nickel		50	49.7	99.4	90 - 110	
Silver		50	49.3	98.6	90 - 110	
Thallium		50	49.1	98.2	90 - 110	
Vanadium		50	49.7	99.5	90 - 110	
Zinc		50	50.7	101	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4829701
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-10
Instrument ID: ICP-MS2
Run Time: 13:34
Method: 6020A
File ID: NI.062816.133459
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0492	mg/L	98.4	90 - 110		
Arsenic		0.0500	0.0495	mg/L	99.0	90 - 110		
Barium		0.0500	0.0487	mg/L	97.5	90 - 110		
Cadmium		0.0500	0.0487	mg/L	97.4	90 - 110		
Chromium		0.0500	0.0494	mg/L	98.8	90 - 110		
Cobalt		0.0500	0.0512	mg/L	102	90 - 110		
Copper		0.0500	0.0496	mg/L	99.2	90 - 110		
Lead		0.0500	0.0496	mg/L	99.3	90 - 110		
Manganese		0.0500	0.0497	mg/L	99.5	90 - 110		
Nickel		0.0500	0.0495	mg/L	98.9	90 - 110		
Silver		0.0500	0.0487	mg/L	97.3	90 - 110		
Thallium		0.0500	0.0489	mg/L	97.7	90 - 110		
Vanadium		0.0500	0.0499	mg/L	99.8	90 - 110		
Zinc		0.0500	0.0495	mg/L	99.0	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829704
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-14
Instrument ID: ICP-MS2
Run Time: 14:31
Method: 6020A
File ID: NI.062816.143138
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0480	mg/L	96.1	90 - 110		
Arsenic		0.0500	0.0501	mg/L	100	90 - 110		
Barium		0.0500	0.0474	mg/L	94.9	90 - 110		
Cadmium		0.0500	0.0480	mg/L	96.1	90 - 110		
Chromium		0.0500	0.0496	mg/L	99.1	90 - 110		
Cobalt		0.0500	0.0531	mg/L	106	90 - 110		
Copper		0.0500	0.0499	mg/L	99.7	90 - 110		
Lead		0.0500	0.0506	mg/L	101	90 - 110		
Manganese		0.0500	0.0492	mg/L	98.5	90 - 110		
Nickel		0.0500	0.0495	mg/L	99.0	90 - 110		
Silver		0.0500	0.0484	mg/L	96.8	90 - 110		
Thallium		0.0500	0.0490	mg/L	98.1	90 - 110		
Vanadium		0.0500	0.0508	mg/L	102	90 - 110		
Zinc		0.0500	0.0501	mg/L	100	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829704
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-18
Instrument ID: ICP-MS2
Run Time: 15:10
Method: 6020A
File ID: NI.062816.151001
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0504	mg/L	101	90 - 110		
Arsenic		0.0500	0.0499	mg/L	99.7	90 - 110		
Barium		0.0500	0.0485	mg/L	96.9	90 - 110		
Cadmium		0.0500	0.0487	mg/L	97.5	90 - 110		
Chromium		0.0500	0.0489	mg/L	97.9	90 - 110		
Cobalt		0.0500	0.0517	mg/L	103	90 - 110		
Copper		0.0500	0.0490	mg/L	97.9	90 - 110		
Lead		0.0500	0.0514	mg/L	103	90 - 110		
Manganese		0.0500	0.0487	mg/L	97.5	90 - 110		
Nickel		0.0500	0.0484	mg/L	96.8	90 - 110		
Silver		0.0500	0.0495	mg/L	98.9	90 - 110		
Thallium		0.0500	0.0506	mg/L	101	90 - 110		
Vanadium		0.0500	0.0498	mg/L	99.6	90 - 110		
Zinc		0.0500	0.0494	mg/L	98.9	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829704
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822 Run Date: 06/28/2016 Sample ID: WG574537-29
Instrument ID: ICP-MS2 Run Time: 18:49 Method: 6020A
File ID: NI.062816.184946 Analyst: BKT QC Key: DOD4
Workgroup (AAB#): WG574570 Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0481	mg/L	96.2	90 - 110		
Arsenic		0.0500	0.0471	mg/L	94.1	90 - 110		
Barium		0.0500	0.0464	mg/L	92.9	90 - 110		
Cadmium		0.0500	0.0463	mg/L	92.7	90 - 110		
Chromium		0.0500	0.0469	mg/L	93.8	90 - 110		
Cobalt		0.0500	0.0469	mg/L	93.8	90 - 110		
Copper		0.0500	0.0475	mg/L	94.9	90 - 110		
Lead		0.0500	0.0484	mg/L	96.7	90 - 110		
Manganese		0.0500	0.0473	mg/L	94.6	90 - 110		
Nickel		0.0500	0.0476	mg/L	95.1	90 - 110		
Silver		0.0500	0.0474	mg/L	94.8	90 - 110		
Thallium		0.0500	0.0474	mg/L	94.8	90 - 110		
Vanadium		0.0500	0.0472	mg/L	94.5	90 - 110		
Zinc		0.0500	0.0472	mg/L	94.5	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829704
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-32
Instrument ID: ICP-MS2
Run Time: 20:23
Method: 6020A
File ID: NI.062816.202325
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0502	mg/L	100	90 - 110		
Arsenic		0.0500	0.0499	mg/L	99.8	90 - 110		
Barium		0.0500	0.0488	mg/L	97.7	90 - 110		
Cadmium		0.0500	0.0494	mg/L	98.8	90 - 110		
Chromium		0.0500	0.0508	mg/L	102	90 - 110		
Cobalt		0.0500	0.0517	mg/L	103	90 - 110		
Copper		0.0500	0.0507	mg/L	101	90 - 110		
Lead		0.0500	0.0515	mg/L	103	90 - 110		
Manganese		0.0500	0.0509	mg/L	102	90 - 110		
Nickel		0.0500	0.0501	mg/L	100	90 - 110		
Silver		0.0500	0.0500	mg/L	99.9	90 - 110		
Thallium		0.0500	0.0504	mg/L	101	90 - 110		
Vanadium		0.0500	0.0505	mg/L	101	90 - 110		
Zinc		0.0500	0.0506	mg/L	101	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829704
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-34
Instrument ID: ICP-MS2
Run Time: 21:01
Method: 6020A
File ID: NI.062816.210146
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0494	mg/L	98.8	90 - 110		
Arsenic		0.0500	0.0505	mg/L	101	90 - 110		
Barium		0.0500	0.0490	mg/L	98.1	90 - 110		
Cadmium		0.0500	0.0492	mg/L	98.4	90 - 110		
Chromium		0.0500	0.0506	mg/L	101	90 - 110		
Cobalt		0.0500	0.0511	mg/L	102	90 - 110		
Copper		0.0500	0.0509	mg/L	102	90 - 110		
Lead		0.0500	0.0534	mg/L	107	90 - 110		
Manganese		0.0500	0.0526	mg/L	105	90 - 110		
Nickel		0.0500	0.0509	mg/L	102	90 - 110		
Silver		0.0500	0.0485	mg/L	96.9	90 - 110		
Thallium		0.0500	0.0506	mg/L	101	90 - 110		
Vanadium		0.0500	0.0505	mg/L	101	90 - 110		
Zinc		0.0500	0.0506	mg/L	101	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829704
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822
Run Date: 06/29/2016
Sample ID: WG574676-10
Instrument ID: ICP-MS2
Run Time: 12:00
Method: 6020A
File ID: NI.062916.120010
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0508	mg/L	102	90 - 110		
Arsenic		0.0500	0.0514	mg/L	103	90 - 110		
Barium		0.0500	0.0505	mg/L	101	90 - 110		
Cadmium		0.0500	0.0516	mg/L	103	90 - 110		
Chromium		0.0500	0.0504	mg/L	101	90 - 110		
Cobalt		0.0500	0.0538	mg/L	108	90 - 110		
Copper		0.0500	0.0507	mg/L	101	90 - 110		
Lead		0.0500	0.0529	mg/L	106	90 - 110		
Manganese		0.0500	0.0534	mg/L	107	90 - 110		
Nickel		0.0500	0.0502	mg/L	100	90 - 110		
Silver		0.0500	0.0508	mg/L	102	90 - 110		
Thallium		0.0500	0.0519	mg/L	104	90 - 110		
Vanadium		0.0500	0.0516	mg/L	103	90 - 110		
Zinc		0.0500	0.0510	mg/L	102	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829704
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822
Run Date: 06/29/2016
Sample ID: WG574676-12
Instrument ID: ICP-MS2
Run Time: 12:37
Method: 6020A
File ID: NI.062916.123748
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0505	mg/L	101	90 - 110		
Arsenic		0.0500	0.0498	mg/L	99.6	90 - 110		
Barium		0.0500	0.0499	mg/L	99.8	90 - 110		
Cadmium		0.0500	0.0506	mg/L	101	90 - 110		
Chromium		0.0500	0.0489	mg/L	97.8	90 - 110		
Cobalt		0.0500	0.0506	mg/L	101	90 - 110		
Copper		0.0500	0.0500	mg/L	99.9	90 - 110		
Lead		0.0500	0.0499	mg/L	99.7	90 - 110		
Manganese		0.0500	0.0500	mg/L	99.9	90 - 110		
Nickel		0.0500	0.0491	mg/L	98.2	90 - 110		
Silver		0.0500	0.0505	mg/L	101	90 - 110		
Thallium		0.0500	0.0497	mg/L	99.4	90 - 110		
Vanadium		0.0500	0.0503	mg/L	101	90 - 110		
Zinc		0.0500	0.0498	mg/L	99.7	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829704
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-07
Instrument ID: ICP-MS2
Run Time: 13:25
Method: 6020A
File ID: NI.062816.132523
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.414	ug/L	104	70 - 130		
Arsenic		0.400	0.419	ug/L	105	70 - 130		
Barium		0.750	0.681	ug/L	90.8	70 - 130		
Cadmium		0.240	0.234	ug/L	97.7	70 - 130		
Chromium		0.800	0.628	ug/L	78.5	70 - 130		
Cobalt		0.400	0.384	ug/L	96.1	70 - 130		
Copper		0.800	0.759	ug/L	94.9	70 - 130		
Lead		0.200	0.183	ug/L	91.6	70 - 130		
Manganese		0.500	0.437	ug/L	87.4	70 - 130		
Nickel		1.60	1.53	ug/L	95.5	70 - 130		
Silver		0.400	0.372	ug/L	92.9	70 - 130		
Thallium		0.0800	0.0768	ug/L	96.0	70 - 130		
Vanadium		0.400	0.332	ug/L	83.0	70 - 130		
Zinc		6.25	6.14	ug/L	98.3	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829706
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-20
Instrument ID: ICP-MS2
Run Time: 15:16
Method: 6020A
File ID: NI.062816.151629
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.429	ug/L	107	70 - 130		
Arsenic		0.400	0.377	ug/L	94.2	70 - 130		
Barium		0.750	0.677	ug/L	90.2	70 - 130		
Cadmium		0.240	0.235	ug/L	98.1	70 - 130		
Chromium		0.800	0.595	ug/L	74.4	70 - 130		
Cobalt		0.400	0.388	ug/L	97.1	70 - 130		
Copper		0.800	0.739	ug/L	92.4	70 - 130		
Lead		0.200	0.192	ug/L	96.0	70 - 130		
Manganese		0.500	0.408	ug/L	81.6	70 - 130		
Nickel		1.60	1.48	ug/L	92.6	70 - 130		
Silver		0.400	0.384	ug/L	96.0	70 - 130		
Thallium		0.0800	0.0804	ug/L	101	70 - 130		
Vanadium		0.400	0.327	ug/L	81.9	70 - 130		
Zinc		6.25	6.24	ug/L	99.9	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829706
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-31
Instrument ID: ICP-MS2
Run Time: 18:56
Method: 6020A
File ID: NI.062816.185610
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.386	ug/L	96.6	70 - 130		
Arsenic		0.400	0.381	ug/L	95.2	70 - 130		
Barium		0.750	0.671	ug/L	89.5	70 - 130		
Cadmium		0.240	0.217	ug/L	90.2	70 - 130		
Chromium		0.800	0.581	ug/L	72.7	70 - 130		
Cobalt		0.400	0.365	ug/L	91.2	70 - 130		
Copper		0.800	0.726	ug/L	90.7	70 - 130		
Lead		0.200	0.188	ug/L	93.9	70 - 130		
Manganese		0.500	0.396	ug/L	79.2	70 - 130		
Nickel		1.60	1.44	ug/L	89.7	70 - 130		
Silver		0.400	0.365	ug/L	91.2	70 - 130		
Thallium		0.0800	0.0752	ug/L	94.0	70 - 130		
Vanadium		0.400	0.309	ug/L	77.3	70 - 130		
Zinc		6.25	5.67	ug/L	90.8	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829706
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060822
Run Date: 06/28/2016
Sample ID: WG574537-44
Instrument ID: ICP-MS2
Run Time: 23:22
Method: 6020A
File ID: NI.062816.232221
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.384	ug/L	96.1	70 - 130		
Arsenic		0.400	0.418	ug/L	105	70 - 130		
Barium		0.750	0.696	ug/L	92.8	70 - 130		
Cadmium		0.240	0.223	ug/L	92.9	70 - 130		
Chromium		0.800	0.654	ug/L	81.8	70 - 130		
Cobalt		0.400	0.379	ug/L	94.6	70 - 130		
Copper		0.800	0.746	ug/L	93.2	70 - 130		
Lead		0.200	0.200	ug/L	100	70 - 130		
Manganese		0.500	0.444	ug/L	88.8	70 - 130		
Nickel		1.60	1.53	ug/L	95.6	70 - 130		
Silver		0.400	0.372	ug/L	93.0	70 - 130		
Thallium		0.0800	0.0774	ug/L	96.8	70 - 130		
Vanadium		0.400	0.335	ug/L	83.8	70 - 130		
Zinc		6.25	5.99	ug/L	95.8	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829706
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060822
Run Date: 06/29/2016
Sample ID: WG574676-07
Instrument ID: ICP-MS2
Run Time: 11:46
Method: 6020A
File ID: NI.062916.114653
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.456	ug/L	114	70 - 130		
Arsenic		0.400	0.429	ug/L	107	70 - 130		
Barium		0.750	0.733	ug/L	97.8	70 - 130		
Cadmium		0.240	0.237	ug/L	98.5	70 - 130		
Chromium		0.800	0.739	ug/L	92.4	70 - 130		
Cobalt		0.400	0.414	ug/L	103	70 - 130		
Copper		0.800	0.768	ug/L	96.0	70 - 130		
Lead		0.200	0.181	ug/L	90.3	70 - 130		
Manganese		0.500	0.468	ug/L	93.6	70 - 130		
Nickel		1.60	1.56	ug/L	97.7	70 - 130		
Silver		0.400	0.390	ug/L	97.5	70 - 130		
Thallium		0.0800	0.0871	ug/L	109	70 - 130		
Vanadium		0.400	0.382	ug/L	95.5	70 - 130		
Zinc		6.25	6.64	ug/L	106	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829706
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060822
Run Date: 06/29/2016
Sample ID: WG574676-14
Instrument ID: ICP-MS2
Run Time: 12:44
Method: 6020A
File ID: NI.062916.124412
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.401	ug/L	100	70 - 130		
Arsenic		0.400	0.394	ug/L	98.6	70 - 130		
Barium		0.750	0.718	ug/L	95.7	70 - 130		
Cadmium		0.240	0.227	ug/L	94.7	70 - 130		
Chromium		0.800	0.768	ug/L	96.0	70 - 130		
Cobalt		0.400	0.405	ug/L	101	70 - 130		
Copper		0.800	0.751	ug/L	93.9	70 - 130		
Lead		0.200	0.179	ug/L	89.4	70 - 130		
Manganese		0.500	0.469	ug/L	93.8	70 - 130		
Nickel		1.60	1.53	ug/L	95.6	70 - 130		
Silver		0.400	0.383	ug/L	95.7	70 - 130		
Thallium		0.0800	0.0809	ug/L	101	70 - 130		
Vanadium		0.400	0.370	ug/L	92.4	70 - 130		
Zinc		6.25	6.89	ug/L	110	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829706
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16060822
Instrument ID: ICP-MS2
Sol. A : WG574537-08 **File ID:** NI.062816.132834
Sol. AB : WG574537-09 **File ID:** NI.062816.133145

Workgroup (AAB#): WG574570
Method: 6020A
Units: ug/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Antimony	NS	0.0836	NS	100	94.7	94.7	
Arsenic	NS	-0.0118	NS	100	98.9	98.9	
Barium	NS	0.266	NS	100	94.5	94.5	
Cadmium	NS	-0.106	NS	100	95.6	95.6	
Chromium	NS	-0.0468	NS	100	96.5	96.5	
Cobalt	NS	0.0276	NS	100	97.7	97.7	
Copper	NS	0.231	NS	100	96.5	96.5	
Lead	NS	0.0352	NS	100	98.2	98.2	
Manganese	NS	0.0680	NS	100	97.7	97.7	
Nickel	NS	0.332	NS	100	96.0	96.0	
Silver	NS	0.0254	NS	100	89.5	89.5	
Thallium	NS	0.0257	NS	100	96.4	96.4	
Vanadium	NS	-0.0466	NS	100	96.5	96.5	
Zinc	NS	0.903	NS	100	101	101	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4829703
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16060822
Instrument ID: ICP-MS2
Sol. A : WG574537-16 **File ID:** NI.062816.150336
Sol. AB : WG574537-17 **File ID:** NI.062816.150647

Workgroup (AAB#): WG574570
Method: 6020A
Units: ug/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Antimony	NS	0.126	NS	100	101	101	
Arsenic	NS	-0.0170	NS	100	102	102	
Barium	NS	0.271	NS	100	98.3	98.3	
Cadmium	NS	-0.0956	NS	100	99.8	99.8	
Chromium	NS	-0.00590	NS	100	97.3	97.3	
Cobalt	NS	0.0242	NS	100	102	102	
Copper	NS	0.234	NS	100	97.4	97.4	
Lead	NS	0.0422	NS	100	102	102	
Manganese	NS	0.0683	NS	100	97.7	97.7	
Nickel	NS	0.343	NS	100	95.9	95.9	
Silver	NS	0.0228	NS	100	91.0	91.0	
Thallium	NS	0.0182	NS	100	99.2	99.2	
Vanadium	NS	-0.0922	NS	100	100	100	
Zinc	NS	0.968	NS	100	103	103	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4829703
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16060822
Instrument ID: ICP-MS2
Sol. A : WG574676-08 **File ID:** NI.062916.115053
Sol. AB : WG574676-09 **File ID:** NI.062916.115447

Workgroup (AAB#): WG574570
Method: 6020A
Units: ug/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Antimony	NS	0.0658	NS	100	100	100	
Arsenic	NS	0.0161	NS	100	105	105	
Barium	NS	0.303	NS	100	98.4	98.4	
Cadmium	NS	-0.103	NS	100	101	101	
Chromium	NS	0.0680	NS	100	101	101	
Cobalt	NS	0.0292	NS	100	105	105	
Copper	NS	0.245	NS	100	100	100	
Lead	NS	0.0174	NS	100	102	102	
Manganese	NS	0.0629	NS	100	105	105	
Nickel	NS	0.331	NS	100	100	100	
Silver	NS	0.0271	NS	100	95.3	95.3	
Thallium	NS	0.0271	NS	100	100	100	
Vanadium	NS	0.0706	NS	100	103	103	
Zinc	NS	1.14	NS	100	104	104	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4829703
Report generated 06/30/2016 09:23

Microbac Laboratories Inc.

INTERNAL STANDARD REPORT

Login: L16060822 **Analytical Method:** 6020
Analytical Workgroup: WG574570 **Matrix:** 1
Instrument: ICP-MS2 **Analyst:** BKT
ICAL Date: 28-JUN-2016 13:06

			BISMUTH	GERMANIUM	INDIUM
Sample	Type	Run Date	% Rec	% Rec	% Rec
L16060822-02	SAMP	28-JUN-2016 20:42	96.563	98.195	99.724
L16060822-03	SAMP	28-JUN-2016 20:45	99.671	105.218	101.024
WG574242-02	BLANK	28-JUN-2016 19:51	99.476	99.537	97.857
WG574242-03	LCS	28-JUN-2016 19:54	99.956	99.501	98.669
WG574537-05	ICV	28-JUN-2016 13:18	106.21	104.263	104.758
WG574537-06	ICB	28-JUN-2016 13:22	103.902	104.951	103.59
WG574537-07	LLICV	28-JUN-2016 13:25	104.131	100.167	102.925
WG574537-08	ICS	28-JUN-2016 13:28	105.453	105.812	102.55
WG574537-09	ICS	28-JUN-2016 13:31	104.707	103.975	104.833
WG574537-10	CCV	28-JUN-2016 13:34	105.977	103.288	105.534
WG574537-11	CCB	28-JUN-2016 13:38	102.403	102.67	102.767
WG574537-14	CCV	28-JUN-2016 14:31	103.407	95.988	101.793
WG574537-15	CCB	28-JUN-2016 14:34	104.225	98.805	100.724
WG574537-16	ICS	28-JUN-2016 15:03	99.952	99.435	100.043
WG574537-17	ICS	28-JUN-2016 15:06	101.575	99.874	97.332
WG574537-18	CCV	28-JUN-2016 15:10	100.244	98.973	99.201
WG574537-19	CCB	28-JUN-2016 15:13	103.776	99.505	101.81
WG574537-20	LLCCV	28-JUN-2016 15:16	99.051	96.111	96.093
WG574537-29	CCV	28-JUN-2016 18:49	100.019	101.137	98.931
WG574537-30	CCB	28-JUN-2016 18:52	100.148	99.202	100.624
WG574537-31	LLCCV	28-JUN-2016 18:56	95.839	97.87	93.931
WG574537-32	CCV	28-JUN-2016 20:23	102.351	100.903	101.064
WG574537-33	CCB	28-JUN-2016 20:26	99.219	96.376	100.773
WG574537-34	CCV	28-JUN-2016 21:01	96.559	99.919	99.677
WG574537-35	CCB	28-JUN-2016 21:04	100.258	102.056	98.569
WG574537-44	LLCCV	28-JUN-2016 23:22	91.439	93.342	91.269
WG574570-01	PSPK	28-JUN-2016 20:13	100.708	100.256	100.218
WG574570-02	SERIAL	28-JUN-2016 20:17	103.535	103.697	101.2

Acceptance criteria: 30% - 120% Underlined recoveries are out of range
Acceptance criteria for CCVs and CCBs for method SW846-6020: 80% - 120%

INT_STD_ICPMS - Modified 07/28/2010
PDF File ID: 4829699
Report generated: 06/30/2016 09:23

Microbac Laboratories Inc.

INTERNAL STANDARD REPORT

Login: L16060822 **Analytical Method:** 6020
Analytical Workgroup: WG574570 **Matrix:** 1
Instrument: ICP-MS2 **Analyst:** BKT
ICAL Date: 29-JUN-2016 11:25

			BISMUTH	GERMANIUM	INDIUM
Sample	Type	Run Date	% Rec	% Rec	% Rec
L16060822-03	SAMP	29-JUN-2016 12:25	99.273	98.877	99.813
WG574676-05	ICV	29-JUN-2016 11:38	98.257	100.661	96.465
WG574676-06	ICB	29-JUN-2016 11:41	98.914	99.091	98.14
WG574676-07	LLICV	29-JUN-2016 11:46	95.79	95.394	93.877
WG574676-08	ICS	29-JUN-2016 11:50	98.596	98.274	96.165
WG574676-09	ICS	29-JUN-2016 11:54	97.891	98.362	95.991
WG574676-10	CCV	29-JUN-2016 12:00	96.559	97.448	96.281
WG574676-11	CCB	29-JUN-2016 12:07	98.969	97.431	96.855
WG574676-12	CCV	29-JUN-2016 12:37	99.303	102.045	96.376
WG574676-13	CCB	29-JUN-2016 12:40	100.991	102.01	98.548
WG574676-14	LLCCV	29-JUN-2016 12:44	96.535	95.69	95.451

Acceptance criteria: 30% - 120% Underlined recoveries are out of range
Acceptance criteria for CCVs and CCBs for method SW846-6020: 80% - 120%

INT_STD_ICPMS - Modified 07/28/2010
PDF File ID: 4829699
Report generated: 06/30/2016 09:23

Microbac Laboratories Inc.

INTERNAL STANDARD REPORT

Login: L16060822 **Analytical Method:** 6020
Analytical Workgroup: WG574570 **Matrix:** 2
Instrument: ICP-MS2 **Analyst:** BKT
ICAL Date: 28-JUN-2016 13:06

			BISMUTH	GERMANIUM	INDIUM
Sample	Type	Run Date	% Rec	% Rec	% Rec
L16060843-01	SAMP	28-JUN-2016 20:10	97.785	95.068	99.061

Acceptance criteria: 30% - 120% Underlined recoveries are out of range
Acceptance criteria for CCVs and CCBs for method SW846-6020: 80% - 120%

INT_STD_ICPMS - Modified 07/28/2010
PDF File ID: 4829699
Report generated: 06/30/2016 09:23

Microbac Laboratories Inc.
LINEAR RANGE (QUARTERLY)

Login Number: L16060822 **Date:** 04/19/2016
Insturment ID: ICP-MS2 **Method:** 6020A

Analyte	Integration Time (Sec.)	Concentration (ug/L)
Antimony	1.00	100.0
Arsenic	1.00	100.0
Barium	1.00	100.0
Cadmium	1.00	100.0
Chromium	1.00	100.0
Cobalt	1.00	100.0
Copper	1.00	100.0
Lead	1.00	100.0
Manganese	1.00	100.0
Nickel	1.00	100.0
Selenium	1.00	100.0
Silver	1.00	100.0
Thallium	1.00	100.0
Uranium	1.00	100.0
Vanadium	1.00	100.0
Zinc	1.00	100.0

Comments:
All analytes passed acceptance criteria at the specified concentration.

LINEAR_RANGE - Modified 03/06/2008
PDF File ID: 4829698
Report generated: 06/29/2016 08:34

2.3.2.3 Raw Data

MassCal File Name

Mass Calibration File Name Default.tun
MassCal File Path C:\NexIONData\MassCal\Default.tun
Peak Search Window: 1.00

Sample Information

Sample Date/Time: Tuesday, June 28, 2016 12:45:36

Mass Calibration and Resolution

Analyte	E Mass	Meas Mass	Mass C DAC Val	Res DAC Value	Meas Peak W	Custom Res
Li	7.016	7.025	1331	2024	0.706	
Mg	23.985	24.025	4508	2020	0.701	
Co	58.933	58.925	11686	2022	0.701	
In	114.904	114.875	22861	2028	0.696	
U	238.050	238.025	47455	2043	0.688	

Relative Std. Dev.

Mass	Meas. Intens. RSD
5.525	5.573
5.575	2.596
5.625	3.076
5.675	2.368
5.725	2.568
5.775	3.011
5.825	2.517
5.875	2.694
5.925	1.595
5.975	3.373
6.025	3.068
6.075	2.439
6.125	2.845
6.175	3.293
6.225	6.959
6.275	54.149
6.325	79.756
6.375	69.722
6.425	99.103
6.475	35.972
6.525	19.569
6.575	4.348
6.625	4.647
6.675	4.116
6.725	2.239
6.775	2.849
6.825	1.600

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 1

Approved: June 29, 2016

6.875	3.248
6.925	2.794
6.975	1.665
7.025	2.844
7.075	2.994
7.125	3.503
7.175	2.980
7.225	3.713
7.275	5.087
7.325	35.264
7.375	93.541
7.425	136.931
7.475	82.402
7.525	71.261
7.575	39.123
7.625	180.662
7.675	64.358
7.725	47.507
7.775	71.261
7.825	70.711
7.875	52.705
7.925	46.481
7.975	63.888
8.025	74.689
8.075	64.359
8.125	63.888
8.175	34.401
8.225	46.481
8.275	49.793
8.325	81.441
8.375	60.858
8.425	67.420
8.475	108.653
22.525	223.607
22.575	90.014
22.625	44.605
22.675	65.617
22.725	58.975
22.775	40.945
22.825	35.460
22.875	18.054
22.925	42.061
22.975	53.715
23.025	19.357
23.075	24.491
23.125	8.438
23.175	12.837

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 2

Approved: June 29, 2016

23.225	30.858
23.275	24.845
23.325	45.141
23.375	26.438
23.425	14.816
23.475	14.263
23.525	12.137
23.575	3.832
23.625	2.363
23.675	1.595
23.725	2.170
23.775	2.317
23.825	1.497
23.875	1.501
23.925	1.731
23.975	0.863
24.025	2.634
24.075	2.734
24.125	2.053
24.175	2.398
24.225	2.993
24.275	2.883
24.325	7.686
24.375	28.384
24.425	36.617
24.475	12.141
24.525	3.189
24.575	0.926
24.625	1.585
24.675	1.547
24.725	2.051
24.775	1.408
24.825	2.678
24.875	1.711
24.925	1.949
24.975	2.595
25.025	3.005
25.075	2.962
25.125	2.366
25.175	2.463
25.225	2.722
25.275	9.971
25.325	31.599
25.375	47.140
25.425	64.885
25.475	6.734
57.525	8.866

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 3

Approved: June 29, 2016

57.575	5.291
57.625	3.440
57.675	3.697
57.725	3.076
57.775	2.567
57.825	2.416
57.875	1.794
57.925	2.766
57.975	2.199
58.025	1.926
58.075	2.254
58.125	1.436
58.175	1.390
58.225	3.504
58.275	5.534
58.325	34.267
58.375	33.055
58.425	20.772
58.475	14.356
58.525	5.854
58.575	3.115
58.625	4.057
58.675	2.315
58.725	2.604
58.775	2.171
58.825	1.290
58.875	1.298
58.925	2.935
58.975	1.929
59.025	1.078
59.075	1.921
59.125	2.163
59.175	2.388
59.225	3.289
59.275	8.929
59.325	38.030
59.375	58.736
59.425	28.284
59.475	17.563
59.525	12.128
59.575	6.526
59.625	2.279
59.675	2.377
59.725	6.247
59.775	4.326
59.825	2.635
59.875	3.246

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 4

Approved: June 29, 2016

59.925	1.954
59.975	3.753
60.025	5.882
60.075	3.380
60.125	2.788
60.175	3.075
60.225	4.430
60.275	18.947
60.325	43.241
60.375	22.017
60.425	79.756
60.475	50.461
113.525	7.111
113.575	1.614
113.625	3.295
113.675	3.378
113.725	2.143
113.775	2.366
113.825	3.191
113.875	4.581
113.925	1.498
113.975	2.463
114.025	3.675
114.075	2.767
114.125	4.064
114.175	2.276
114.225	6.994
114.275	28.008
114.325	41.691
114.375	17.526
114.425	9.127
114.475	6.547
114.525	3.985
114.575	2.757
114.625	3.785
114.675	2.945
114.725	3.014
114.775	2.252
114.825	1.319
114.875	1.984
114.925	1.415
114.975	0.577
115.025	0.422
115.075	1.851
115.125	0.695
115.175	3.426
115.225	3.732

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 5

Approved: June 29, 2016

115.275	9.818
115.325	57.810
115.375	52.705
115.425	23.458
115.475	18.657
115.525	4.767
115.575	6.521
115.625	4.984
115.675	3.649
115.725	7.463
115.775	7.302
115.825	2.294
115.875	3.587
115.925	2.925
115.975	3.587
116.025	4.548
116.075	6.187
116.125	2.246
116.175	11.560
116.225	3.389
116.275	17.538
116.325	51.618
116.375	52.705
116.425	42.127
116.475	34.312
236.525	
236.575	28.853
236.625	56.165
236.675	51.701
236.725	11.620
236.775	50.775
236.825	27.217
236.875	33.431
236.925	29.881
236.975	45.955
237.025	46.771
237.075	7.402
237.125	30.162
237.175	24.686
237.225	34.297
237.275	16.110
237.325	20.328
237.375	28.770
237.425	26.702
237.475	18.887
237.525	24.816
237.575	6.725

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 6

Approved: June 29, 2016

237.625	8.740
237.675	3.668
237.725	2.023
237.775	1.587
237.825	3.110
237.875	1.530
237.925	2.794
237.975	3.068
238.025	2.967
238.075	2.352
238.125	2.652
238.175	2.474
238.225	2.129
238.275	1.500
238.325	2.197
238.375	3.421
238.425	3.111
238.475	6.669
238.525	16.077
238.575	17.557
238.625	21.969
238.675	32.775
238.725	22.232
238.775	26.146
238.825	21.694
238.875	19.088
238.925	8.384
238.975	17.180
239.025	59.843
239.075	31.485
239.125	33.497
239.175	31.181
239.225	27.834
239.275	28.743
239.325	26.503
239.375	17.174
239.425	9.195
239.475	35.446

Approved: June 29, 2016

SmartTune Wizard - Summary

```
Optimization Summary

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Start Time: 6/28/2016 12:50:43 PM
End Time: 6/28/2016 12:52:55 PM

Daily Performance Check - [Passed]  Optimum value(s): N/A
        Obtained Intensity (Be 9.0122): 5221.30
        Obtained Intensity (Mg 23.985): 136183.87
        Obtained Intensity (In 114.904): 66441.11
        Obtained Intensity (U 238.05): 147068.09
        Obtained Intensity (Bkgd 220): 0.20
        Obtained Formula (CeO 155.9 / Ce 139.905): 0.018 (=3731.53 / 205818.69)
        Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.004 (=890.23 / 205818.69)
```

Report Date/Time: Tuesday, June 28, 2016 12:52:56
Page 1

Approved: June 29, 2016

SmartTune Wizard - Details

```
Optimization Details

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Optimization Status

Start Time: 6/28/2016 12:50:43 PM

Daily Performance Check
        Optimization Settings:
                Method: C:\NexIONData\Method\ESI Daily Performance.mth.
                Intensity Criterion: Be 9.0122 > 2000
                Intensity Criterion: Mg 23.985 > 15000
                Intensity Criterion: In 114.904 > 40000
                Intensity Criterion: U 238.05 > 30000
                Intensity Criterion: Bkgd 220 <= 5
                Formula Criterion: CeO 155.9 / Ce 139.905 <= 0.025
                Formula Criterion: Ce++ 69.9527 / Ce 139.905 <= 0.03

        Optimization Results:
        Initial Try
                Obtained Intensity (Be 9.0122): 5221.30
                Obtained Intensity (Mg 23.985): 136183.87
                Obtained Intensity (In 114.904): 66441.11
                Obtained Intensity (U 238.05): 147068.09
                Obtained Intensity (Bkgd 220): 0.20
                Obtained Formula (CeO 155.9 / Ce 139.905): 0.018 (=3731.53 / 205818.69)
                Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.004 (=890.23 / 205818.69)
[Passed]  Optimum value(s): N/A

End Time: 6/28/2016 12:52:55 PM
```


Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: Blank
Sample Date/Time: Tuesday, June 28, 2016 13:02:57
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	81760.1	2.0				ug/L		Standard
	Be	9	16.7	75.5				ug/L		Standard
	Al	27	1410.1	2.3				ug/L		Standard
	Sc	45	39543.2	2.7				ug/L		Standard
	Ti	47	39.0	9.2				ug/L		Standard
	V	51	1415.2	6.8				ug/L		Standard
	Cr	52	6946.3	1.7				ug/L		Standard
	Cr	53	328.3	14.1				ug/L		Standard
	Mn	55	823.0	8.8				ug/L		Standard
	Co	59	251.7	9.6				ug/L		Standard
	Ni	60	192.0	8.0				ug/L		Standard
	Cu	65	140.7	5.3				ug/L		Standard
	Zn	66	490.7	3.4				ug/L		Standard
>	Ge	72	515086.9	1.2				ug/L		Standard
	As	75	19.8	169.9				ug/L		Standard
	Se	82	15.3	19.8				ug/L		Standard
	Se-1	77	48.0	15.7				ug/L		Standard
>	Ga	71	90.0	5.6				mg/L		Standard
	Rb	85	23.3	68.9				ug/L		Standard
	Y	89	448509.9	0.5				ug/L		Standard
>	Rh	103	15.0	66.7				ug/L		Standard
	Mo	98	10.9	19.2				ug/L		Standard
	Ag	107	134.0	44.7				ug/L		Standard
	Cd	111	6.6	57.0				mg/L		Standard
	Cd	114	16.1	68.4				ug/L		Standard
>	In	115	611116.5	1.9				ug/L		Standard
	Sn	118	518.3	7.2				ug/L		Standard
	Sb	123	126.7	13.8				ug/L		Standard
	Ba	135	45.3	39.8				ug/L		Standard
	Ce	140	298.3	72.3				ug/L		Standard
>	Tb	159	977942.3	1.0				ug/L		Standard
	Ho	165	11.7	24.7				ug/L		Standard
	Tl	203	30.7	114.0				ug/L		Standard
	Tl	205	20.0	152.1				ug/L		Standard
	Pb	206	285.7	7.2				ug/L		Standard
	Pb	207	257.0	5.7				ug/L		Standard
	Pb	208	955.7	3.9				ug/L		Standard
	U	238	17.0	67.8				ug/L		Standard
>	Bi	209	557080.4	0.4				ug/L		Standard

Sample ID: Blank
Report Date/Time: Tuesday, June 28, 2016 13:05:14
Page 1

Approved: June 29, 2016

	Na	23	0.0		mg/L	Standard
	Mg	24	31.7	36.5	mg/L	Standard
	K	39	30.0	57.7	mg/L	Standard
	Ca	43	48.3	43.1	mg/L	Standard
	Fe	54	159.3	19.3	mg/L	Standard
	Fe	57	195.0	11.2	mg/L	Standard
>	**Sc-1**	45	39543.2	2.7	mg/L	Standard
	Cl	35	27641.9	2.8	ug/L	Standard
	Kr	83	2.0	50.0	ug/L	Standard
	Br	81	1646.8	11.1	ug/L	Standard
	P	31	15231.0	1.6	ug/L	Standard
	S	34	1845.1	7.5	ug/L	Standard
	Sr	88	70.0	12.4	ug/L	Standard
	C	12	113.3	28.4	mg/L	Standard
	N	14	0.0		mg/L	Standard
	Hg	202	0.0		mg/L	Standard
	Dy	164	12.4	122.2	mg/L	Standard
	Ho-1	165	11.7	24.7	mg/L	Standard
	Er	166	20.0	50.0	mg/L	Standard
	I	127	1161.7	12.0	mg/L	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Blank
Report Date/Time: Tuesday, June 28, 2016 13:05:14
Page 2

Approved: June 29, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Blank
Report Date/Time: Tuesday, June 28, 2016 13:05:14
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: Standard 1
Sample Date/Time: Tuesday, June 28, 2016 13:06:09
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83638.6	0.9				ug/L	81760	Standard
	Be	9	11.7	49.5				ug/L	17	Standard
	Al	27	1445.1	4.5				ug/L	1410	Standard
	Sc	45	40184.9	2.0				ug/L	39543	Standard
	Ti	47	38.3	12.0				ug/L	39	Standard
	V	51	1492.6	2.6				ug/L	1415	Standard
	Cr	52	7409.5	2.1				ug/L	6946	Standard
	Cr	53	321.7	15.3				ug/L	328	Standard
	Mn	55	835.0	1.8				ug/L	823	Standard
	Co	59	226.7	11.0				ug/L	252	Standard
	Ni	60	182.7	5.1				ug/L	192	Standard
	Cu	65	134.7	15.9				ug/L	141	Standard
	Zn	66	519.7	4.6				ug/L	491	Standard
>	**Ge**	72	517782.5	2.4				ug/L	515087	Standard
	As	75	-33.8	72.0				ug/L	20	Standard
	Se	82	16.9	27.2				ug/L	15	Standard
	Se-1	77	48.7	14.9				ug/L	48	Standard
>	**Ga**	71	95.0	19.0				mg/L	90	Standard
	Rb	85	33.3	82.6				ug/L	23	Standard
	Y	89	456301.2	1.1				ug/L	448510	Standard
>	**Rh**	103	21.7	58.1				ug/L	15	Standard
	Mo	98	8.3	58.9				ug/L	11	Standard
	Ag	107	79.0	9.1				ug/L	134	Standard
	Cd	111	3.3	17.5				mg/L	7	Standard
	Cd	114	6.2	12.1				ug/L	16	Standard
>	**In**	115	624714.7	2.4				ug/L	611117	Standard
	Sn	118	515.0	5.4				ug/L	518	Standard
	Sb	123	47.7	15.0				ug/L	127	Standard
	Ba	135	33.7	27.6				ug/L	45	Standard
	Ce	140	86.7	24.0				ug/L	298	Standard
>	**Tb**	159	1001347.1	1.8				ug/L	977942	Standard
	Ho	165	6.7	43.3				ug/L	12	Standard
	Tl	203	5.3	60.3				ug/L	31	Standard
	Tl	205	3.3	86.6				ug/L	20	Standard
	Pb	206	276.7	3.7				ug/L	286	Standard
	Pb	207	213.3	7.3				ug/L	257	Standard
	Pb	208	933.0	3.4				ug/L	956	Standard
	U	238	4.7	12.4				ug/L	17	Standard
>	**Bi**	209	562888.8	0.9				ug/L	557080	Standard

Sample ID: Standard 1
Report Date/Time: Tuesday, June 28, 2016 13:08:26
Page 1

Approved: June 29, 2016

	Na	23	0.0		mg/L	0	Standard
	Mg	24	36.7	7.9	mg/L	32	Standard
	K	39	28.3	44.4	mg/L	30	Standard
	Ca	43	36.7	28.4	mg/L	48	Standard
	Fe	54	172.5	32.0	mg/L	159	Standard
	Fe	57	211.7	34.4	mg/L	195	Standard
>	**Sc-1**	45	40184.9	2.0	mg/L	39543	Standard
	Cl	35	29114.6	1.1	ug/L	27642	Standard
	Kr	83	3.3	96.4	ug/L	2	Standard
	Br	81	1676.8	6.8	ug/L	1647	Standard
	P	31	15918.4	1.2	ug/L	15231	Standard
	S	34	1885.1	11.3	ug/L	1845	Standard
	Sr	88	58.3	21.6	ug/L	70	Standard
	C	12	116.7	24.7	mg/L	113	Standard
	N	14	0.0		mg/L	0	Standard
	Hg	202	0.0		mg/L	0	Standard
	Dy	164	16.2	93.4	mg/L	12	Standard
	Ho-1	165	6.7	43.3	mg/L	12	Standard
	Er	166	10.0	100.0	mg/L	20	Standard
	I	127	1170.0	16.3	mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Standard 1
Report Date/Time: Tuesday, June 28, 2016 13:08:26
Page 2

Approved: June 29, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 1
Report Date/Time: Tuesday, June 28, 2016 13:08:26
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: Standard 2
Sample Date/Time: Tuesday, June 28, 2016 13:09:21
Number of Replicates: 3
Autosampler Position: 2
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82626.5	2.9				ug/L	81760	Standard
	Be	9	60.0	14.4				ug/L	17	Standard
	Al	27	10023.3	4.0				ug/L	1410	Standard
	Sc	45	41007.1	1.4				ug/L	39543	Standard
	Ti	47	69.0	5.8				ug/L	39	Standard
	V	51	1868.4	2.4				ug/L	1415	Standard
	Cr	52	8071.5	1.1				ug/L	6946	Standard
	Cr	53	360.0	4.2				ug/L	328	Standard
	Mn	55	1347.7	1.9				ug/L	823	Standard
	Co	59	599.0	5.9				ug/L	252	Standard
	Ni	60	305.7	5.8				ug/L	192	Standard
	Cu	65	280.3	2.3				ug/L	141	Standard
	Zn	66	1008.0	4.7				ug/L	491	Standard
>	**Ge**	72	539572.1	1.8				ug/L	515087	Standard
	As	75	11.1	230.1				ug/L	20	Standard
	Se	82	20.4	17.1				ug/L	15	Standard
	Se-1	77	60.7	1.9				ug/L	48	Standard
>	**Ga**	71	83.3	19.3				mg/L	90	Standard
	Rb	85	31.7	48.2				ug/L	23	Standard
	Y	89	463986.2	1.2				ug/L	448510	Standard
>	**Rh**	103	13.3	57.3				ug/L	15	Standard
	Mo	98	329.5	2.8				ug/L	11	Standard
	Ag	107	455.0	1.5				ug/L	134	Standard
	Cd	111	115.7	6.5				mg/L	7	Standard
	Cd	114	292.0	6.0				ug/L	16	Standard
>	**In**	115	625608.8	0.7				ug/L	611117	Standard
	Sn	118	920.0	9.2				ug/L	518	Standard
	Sb	123	303.4	7.6				ug/L	127	Standard
	Ba	135	158.7	8.0				ug/L	45	Standard
	Ce	140	88.3	42.5				ug/L	298	Standard
>	**Tb**	159	1002874.0	1.4				ug/L	977942	Standard
	Ho	165	11.7	24.7				ug/L	12	Standard
	Tl	203	472.7	2.0				ug/L	31	Standard
	Tl	205	391.7	11.0				ug/L	20	Standard
	Pb	206	628.7	5.7				ug/L	286	Standard
	Pb	207	539.3	6.7				ug/L	257	Standard
	Pb	208	2133.7	2.7				ug/L	956	Standard
	U	238	388.7	5.1				ug/L	17	Standard
>	**Bi**	209	570508.1	0.4				ug/L	557080	Standard

Sample ID: Standard 2
Report Date/Time: Tuesday, June 28, 2016 13:11:38
Page 1

Approved: June 29, 2016

	Na	23	0.0		mg/L	0	Standard
	Mg	24	48.3	26.0	mg/L	32	Standard
	K	39	45.0	19.2	mg/L	30	Standard
	Ca	43	36.7	43.8	mg/L	48	Standard
	Fe	54	162.5	17.6	mg/L	159	Standard
	Fe	57	193.3	3.0	mg/L	195	Standard
>	**Sc-1**	45	41007.1	1.4	mg/L	39543	Standard
	Cl	35	29715.1	0.6	ug/L	27642	Standard
	Kr	83	3.0	33.3	ug/L	2	Standard
	Br	81	1836.8	4.6	ug/L	1647	Standard
	P	31	16030.1	3.1	ug/L	15231	Standard
	S	34	1895.1	1.4	ug/L	1845	Standard
	Sr	88	61.7	9.4	ug/L	70	Standard
	C	12	143.3	21.3	mg/L	113	Standard
	N	14	0.0		mg/L	0	Standard
	Hg	202	0.0		mg/L	0	Standard
	Dy	164	19.5	50.0	mg/L	12	Standard
	Ho-1	165	11.7	24.7	mg/L	12	Standard
	Er	166	10.0	100.0	mg/L	20	Standard
	I	127	1208.4	23.1	mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Standard 2
Report Date/Time: Tuesday, June 28, 2016 13:11:38
Page 2

Approved: June 29, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 2
Report Date/Time: Tuesday, June 28, 2016 13:11:38
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: Standard 3
Sample Date/Time: Tuesday, June 28, 2016 13:12:32
Number of Replicates: 3
Autosampler Position: 3
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83151.0	2.6				ug/L	81760	Standard
	Be	9	50594.4	3.3	**50.0000**	0.655	1.3	ug/L	17	Standard
	Al	27	7832644.8	2.3	**50.0000**	1.818	3.6	ug/L	1410	Standard
	Sc	45	42972.5	0.7				ug/L	39543	Standard
	Ti	47	26044.7	0.2	**100.0000**	1.044	1.0	ug/L	39	Standard
	V	51	369925.9	1.5	**50.0000**	0.431	0.9	ug/L	1415	Standard
	Cr	52	376035.2	0.6	**50.0000**	0.381	0.8	ug/L	6946	Standard
	Cr	53	47201.8	2.5	**50.0000**	1.648	3.3	ug/L	328	Standard
	Mn	55	339840.1	0.9	**50.0000**	0.028	0.1	ug/L	823	Standard
	Co	59	369835.8	1.4	**50.0000**	0.329	0.7	ug/L	252	Standard
	Ni	60	100390.8	0.3	**50.0000**	0.410	0.8	ug/L	192	Standard
	Cu	65	101743.1	0.9	**50.0000**	0.203	0.4	ug/L	141	Standard
	Zn	66	59304.5	0.9	**50.0000**	0.371	0.7	ug/L	491	Standard
>	**Ge**	72	547553.1	0.8				ug/L	515087	Standard
	As	75	59263.2	0.8	**50.0000**	0.205	0.4	ug/L	20	Standard
	Se	82	6019.0	1.0	**50.0000**	0.596	1.2	ug/L	15	Standard
	Se-1	77	4068.9	1.3	**50.0000**	0.434	0.9	ug/L	48	Standard
>	**Ga**	71	120.0	22.0				mg/L	90	Standard
	Rb	85	803.4	2.6				ug/L	23	Standard
	Y	89	462085.4	0.0				ug/L	448510	Standard
>	**Rh**	103	33.3	52.7				ug/L	15	Standard
	Mo	98	278482.4	0.5	**100.0000**	0.941	0.9	ug/L	11	Standard
	Ag	107	351728.0	0.3	**50.0000**	0.642	1.3	ug/L	134	Standard
	Cd	111	110396.3	1.4	**50.0000**	0.556	1.1	mg/L	7	Standard
	Cd	114	267604.2	1.4	**50.0000**	0.663	1.3	ug/L	16	Standard
>	**In**	115	627830.2	1.2				ug/L	611117	Standard
	Sn	118	304675.0	1.7	**50.0000**	0.618	1.2	ug/L	518	Standard
	Sb	123	244739.5	0.4	**50.0000**	0.675	1.4	ug/L	127	Standard
	Ba	135	106555.1	1.0	**50.0000**	0.097	0.2	ug/L	45	Standard
	Ce	140	116.7	9.9				ug/L	298	Standard
>	**Tb**	159	1018565.7	2.1				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	438004.0	1.2	**50.0000**	0.427	0.9	ug/L	31	Standard
	Tl	205	379940.6	1.7	**50.0000**	0.597	1.2	ug/L	20	Standard
	Pb	206	284208.1	0.3	**50.0000**	0.335	0.7	ug/L	286	Standard
	Pb	207	256392.2	0.5	**50.0000**	0.374	0.7	ug/L	257	Standard
	Pb	208	978425.0	1.4	**50.0000**	0.588	1.2	ug/L	956	Standard
	U	238	371095.6	0.5	**50.0000**	0.490	1.0	ug/L	17	Standard
>	**Bi**	209	583572.7	0.6				ug/L	557080	Standard

Sample ID: Standard 3
Report Date/Time: Tuesday, June 28, 2016 13:14:49
Page 1

Approved: June 29, 2016

Na	23	0.0					mg/L	0	Standard
Mg	24	4002.2	2.2	**5.0000**	0.144	2.9	mg/L	32	Standard
K	39	1640.1	4.4	**5.0000**	0.220	4.4	mg/L	30	Standard
Ca	43	73.3	27.6	**5.0000**	2.832	56.6	mg/L	48	Standard
Fe	54	6006.1	1.4	**5.0000**	0.034	0.7	mg/L	159	Standard
Fe	57	1818.4	10.7	**5.0000**	0.606	12.1	mg/L	195	Standard
> **Sc-1**	45	42972.5	0.7				mg/L	39543	Standard
Cl	35	31584.9	0.4				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1736.8	12.0				ug/L	1647	Standard
P	31	18696.5	1.2				ug/L	15231	Standard
S	34	2423.5	4.1				ug/L	1845	Standard
Sr	88	71.7	21.3				ug/L	70	Standard
C	12	110.0	39.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	18.6	110.3				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	30.0	33.3				mg/L	20	Standard
I	127	1405.1	8.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72			
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: Standard 3
Report Date/Time: Tuesday, June 28, 2016 13:14:49
Page 2

Approved: June 29, 2016

⎣	Rb	85
⎡	Y	89
⎣>	Rh	103
⎡	Mo	98
⎢	Ag	107
⎢	Cd	111
⎢	Cd	114
⎢>	In	115
⎢	Sn	118
⎢	Sb	123
⎣	Ba	135
⎡	Ce	140
⎣>	Tb	159
⎡	Ho	165
⎢	Tl	203
⎢	Tl	205
⎢	Pb	206
⎢	Pb	207
⎢	Pb	208
⎢	U	238
⎣>	Bi	209
⎡	Na	23
⎢	Mg	24
⎢	K	39
⎢	Ca	43
⎢	Fe	54
⎢	Fe	57
⎣>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 3
Report Date/Time: Tuesday, June 28, 2016 13:14:49
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: Standard 4
Sample Date/Time: Tuesday, June 28, 2016 13:15:44
Number of Replicates: 3
Autosampler Position: 4
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	81919.2	1.8				ug/L	81760	Standard
	Be	9	99366.8	4.2	**99.8402**	3.215	3.2	ug/L	17	Standard
	Al	27	15456145.8	2.8	**100.0514**	1.254	1.3	ug/L	1410	Standard
	Sc	45	41889.5	2.4				ug/L	39543	Standard
	Ti	47	52065.7	3.9	**203.2707**	7.861	3.9	ug/L	39	Standard
	V	51	718696.0	1.3	**100.2818**	2.594	2.6	ug/L	1415	Standard
	Cr	52	727396.2	2.7	**100.4747**	2.418	2.4	ug/L	6946	Standard
	Cr	53	90202.9	3.4	**99.4979**	3.023	3.0	ug/L	328	Standard
	Mn	55	679268.3	1.6	**101.6708**	2.363	2.3	ug/L	823	Standard
	Co	59	730686.9	1.6	**101.0312**	2.535	2.5	ug/L	252	Standard
	Ni	60	193870.4	2.0	**99.9159**	1.756	1.8	ug/L	192	Standard
	Cu	65	194619.8	1.4	**99.4389**	1.906	1.9	ug/L	141	Standard
	Zn	66	113576.2	1.7	**99.8564**	1.521	1.5	ug/L	491	Standard
>	Ge	72	530148.8	1.5				ug/L	515087	Standard
	As	75	116520.9	0.8	**100.7477**	1.604	1.6	ug/L	20	Standard
	Se	82	12083.5	1.4	**101.8724**	2.244	2.2	ug/L	15	Standard
	Se-1	77	7795.3	2.5	**99.8242**	2.540	2.5	ug/L	48	Standard
>	Ga	71	125.0	22.3				mg/L	90	Standard
	Rb	85	1618.4	10.1				ug/L	23	Standard
	Y	89	461192.2	0.4				ug/L	448510	Standard
>	Rh	103	56.7	20.4				ug/L	15	Standard
	Mo	98	556218.8	1.0	**199.4538**	3.118	1.6	ug/L	11	Standard
	Ag	107	678834.5	1.8	**98.0160**	2.327	2.4	ug/L	134	Standard
	Cd	111	219330.5	1.3	**99.4613**	1.808	1.8	mg/L	7	Standard
	Cd	114	534582.7	2.1	**99.7362**	2.615	2.6	ug/L	16	Standard
>	In	115	630478.6	0.6				ug/L	611117	Standard
	Sn	118	609964.7	1.0	**99.8933**	1.520	1.5	ug/L	518	Standard
	Sb	123	482106.0	1.2	**99.0353**	1.710	1.7	ug/L	127	Standard
	Ba	135	212119.9	0.9	**99.5711**	1.444	1.4	ug/L	45	Standard
	Ce	140	165.0	42.7				ug/L	298	Standard
>	Tb	159	1009801.9	0.4				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	865813.4	1.1	**100.7605**	1.348	1.3	ug/L	31	Standard
	Tl	205	752692.6	1.1	**100.8681**	0.657	0.7	ug/L	20	Standard
	Pb	206	549987.8	0.2	**99.7256**	0.843	0.8	ug/L	286	Standard
	Pb	207	495989.2	0.5	**99.7047**	0.496	0.5	ug/L	257	Standard
	Pb	208	1945321.0	0.5	**101.0779**	0.484	0.5	ug/L	956	Standard
	U	238	738847.3	0.4	**101.1182**	1.306	1.3	ug/L	17	Standard
>	Bi	209	568134.9	1.0				ug/L	557080	Standard

Sample ID: Standard 4
Report Date/Time: Tuesday, June 28, 2016 13:18:01
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**19.9850**	17.308	86.6	mg/L	0	Standard
Mg	24	8245.6	0.3	**10.3081**	0.247	2.4	mg/L	32	Standard
K	39	3202.0	9.6	**10.0847**	1.159	11.5	mg/L	30	Standard
Ca	43	108.3	26.2	**10.2160**	4.199	41.1	mg/L	48	Standard
Fe	54	12180.7	3.3	**10.2708**	0.331	3.2	mg/L	159	Standard
Fe	57	3313.7	6.4	**9.9500**	0.928	9.3	mg/L	195	Standard
Sc-1	45	41889.5	2.4				mg/L	39543	Standard
Cl	35	30969.0	3.8				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1773.4	7.0				ug/L	1647	Standard
P	31	19072.0	2.1				ug/L	15231	Standard
S	34	2495.2	5.2				ug/L	1845	Standard
Sr	88	56.7	5.1				ug/L	70	Standard
C	12	143.3	29.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	13.0	113.4				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	6.7	173.2				mg/L	20	Standard
I	127	1551.7	12.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72			
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: Standard 4
Report Date/Time: Tuesday, June 28, 2016 13:18:01
Page 2

Approved: June 29, 2016

	Rb	85
	Y	89
>	Rh	103
	Mo	98
	Ag	107
	Cd	111
	Cd	114
>	In	115
	Sn	118
	Sb	123
	Ba	135
	Ce	140
>	Tb	159
	Ho	165
	Tl	203
	Tl	205
	Pb	206
	Pb	207
	Pb	208
	U	238
>	Bi	209
	Na	23
	Mg	24
	K	39
	Ca	43
	Fe	54
	Fe	57
>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Corr. Coef.	Na	23	Correlation coefficient < 0.998

Sample ID: Standard 4
Report Date/Time: Tuesday, June 28, 2016 13:18:01
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 1
Sample Date/Time: Tuesday, June 28, 2016 13:18:57
Number of Replicates: 3
Autosampler Position: 201
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81910.8	1.2				ug/L	81760	Standard
	Be	9	50188.0	2.5	**50.4285**	0.671	1.3	ug/L	17	Standard
	Al	27	7919939.7	1.3	**51.2716**	0.184	0.4	ug/L	1410	Standard
	Sc	45	42151.9	1.0				ug/L	39543	Standard
	Ti	47	26799.3	0.9	**103.2216**	1.931	1.9	ug/L	39	Standard
	V	51	362002.1	2.3	**49.7477**	0.555	1.1	ug/L	1415	Standard
	Cr	52	370323.8	0.7	**49.9789**	0.741	1.5	ug/L	6946	Standard
	Cr	53	45478.3	5.4	**49.3268**	1.623	3.3	ug/L	328	Standard
	Mn	55	340081.6	0.1	**50.1792**	1.001	2.0	ug/L	823	Standard
	Co	59	367567.5	0.6	**50.1558**	0.780	1.6	ug/L	252	Standard
	Ni	60	97767.5	1.8	**49.6873**	0.150	0.3	ug/L	192	Standard
	Cu	65	99441.7	2.0	**50.1064**	0.391	0.8	ug/L	141	Standard
	Zn	66	58882.6	2.1	**50.6983**	0.019	0.0	ug/L	491	Standard
>	**Ge**	72	537043.0	2.1				ug/L	515087	Standard
	As	75	58926.0	1.4	**50.3136**	0.381	0.8	ug/L	20	Standard
	Se	82	6003.9	1.4	**49.9034**	0.396	0.8	ug/L	15	Standard
	Se-1	77	4037.5	1.3	**50.7056**	1.619	3.2	ug/L	48	Standard
>	**Ga**	71	113.3	22.2				mg/L	90	Standard
	Rb	85	626.7	7.4				ug/L	23	Standard
	Y	89	465817.4	0.7				ug/L	448510	Standard
>	**Rh**	103	43.3	24.0				ug/L	15	Standard
	Mo	98	284447.8	1.6	**100.4857**	3.626	3.6	ug/L	11	Standard
	Ag	107	346501.9	1.0	**49.2778**	1.406	2.9	ug/L	134	Standard
	Cd	111	109630.9	1.1	**48.9766**	1.535	3.1	mg/L	7	Standard
	Cd	114	266600.1	1.2	**48.9905**	1.146	2.3	ug/L	16	Standard
>	**In**	115	640190.8	2.0				ug/L	611117	Standard
	Sn	118	332613.3	1.3	**53.6148**	1.570	2.9	ug/L	518	Standard
	Sb	123	238639.0	2.2	**48.2943**	2.007	4.2	ug/L	127	Standard
	Ba	135	104330.0	1.0	**48.2232**	0.702	1.5	ug/L	45	Standard
	Ce	140	120.0	15.0				ug/L	298	Standard
>	**Tb**	159	1038346.1	1.3				ug/L	977942	Standard
	Ho	165	16.7	34.6				ug/L	12	Standard
	Tl	203	439433.3	1.3	**49.1017**	0.655	1.3	ug/L	31	Standard
	Tl	205	379171.3	1.2	**48.7970**	1.107	2.3	ug/L	20	Standard
	Pb	206	282861.9	1.4	**49.2211**	1.133	2.3	ug/L	286	Standard
	Pb	207	247275.0	1.2	**47.7052**	1.056	2.2	ug/L	257	Standard
	Pb	208	979823.4	0.2	**48.8569**	0.617	1.3	ug/L	956	Standard
	U	238	361651.7	0.7	**47.5263**	0.839	1.8	ug/L	17	Standard
>	**Bi**	209	591677.6	1.0				ug/L	557080	Standard

Sample ID: QC Std 1
Report Date/Time: Tuesday, June 28, 2016 13:21:14
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.0435**	17.387	173.1	mg/L	0	Standard
Mg	24	4277.3	3.5	**5.2835**	0.147	2.8	mg/L	32	Standard
K	39	1651.8	3.6	**5.0931**	0.136	2.7	mg/L	30	Standard
Ca	43	81.7	30.2	**6.2885**	3.508	55.8	mg/L	48	Standard
Fe	54	5996.6	1.6	**4.9548**	0.129	2.6	mg/L	159	Standard
Fe	57	1723.4	6.4	**4.8356**	0.391	8.1	mg/L	195	Standard
Sc-1	45	42151.9	1.0				mg/L	39543	Standard
Cl	35	29861.4	1.9				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1713.4	14.2				ug/L	1647	Standard
P	31	18880.1	1.3				ug/L	15231	Standard
S	34	2316.8	4.8				ug/L	1845	Standard
Sr	88	70.0	35.7				ug/L	70	Standard
C	12	156.7	32.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	46.2	33.1				mg/L	12	Standard
Ho-1	165	16.7	34.6				mg/L	12	Standard
Er	166	10.0					mg/L	20	Standard
I	127	1246.7	11.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	100.857		
Al	27	102.543		
Sc	45			
Ti	47	103.222		
V	51	99.495		
Cr	52	99.958		
Cr	53			
Mn	55	100.358		
Co	59	100.312		
Ni	60	99.375		
Cu	65	100.213		
Zn	66	101.397		
Ge	72		104.263	
As	75	100.627		
Se	82	99.807		
Se-1	77			
Ga	71			

Sample ID: QC Std 1
Report Date/Time: Tuesday, June 28, 2016 13:21:14
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	100.486		
Ag	107	98.556		
Cd	111	97.953		
Cd	114			
> In	115		104.758	
Sn	118	107.230		
Sb	123	96.589		
Ba	135	96.446		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	98.203		
Tl	205			
Pb	206	98.442		
Pb	207	95.410		
Pb	208	97.714		
U	238	95.053		
> Bi	209		106.210	
Na	23	200.869		
Mg	24	105.670		
K	39	101.861		
Ca	43	125.771		
Fe	54	99.097		
Fe	57	96.711		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 1	Na	23	
QC Std 1	Ca	43	

Sample ID: QC Std 1
Report Date/Time: Tuesday, June 28, 2016 13:21:14
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 2
Sample Date/Time: Tuesday, June 28, 2016 13:22:10
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81098.1	0.9				ug/L	81760	Standard
	Be	9	35.0	14.3	**0.0258**	0.005	18.6	ug/L	17	Standard
	Al	27	2568.6	16.6	**0.0024**	0.003	117.4	ug/L	1410	Standard
	Sc	45	40706.3	0.9				ug/L	39543	Standard
	Ti	47	40.7	7.5	**-0.0093**	0.010	110.5	ug/L	39	Standard
	V	51	1184.2	4.7	**-0.0443**	0.005	11.8	ug/L	1415	Standard
	Cr	52	6261.3	1.2	**-0.1999**	0.018	9.1	ug/L	6946	Standard
	Cr	53	338.3	3.7	**0.0258**	0.014	55.9	ug/L	328	Standard
	Mn	55	842.4	4.2	**-0.0247**	0.005	18.3	ug/L	823	Standard
	Co	59	286.0	9.3	**0.0074**	0.003	40.3	ug/L	252	Standard
	Ni	60	199.7	6.5	**-0.0040**	0.005	124.9	ug/L	192	Standard
	Cu	65	142.3	4.7	**-0.0195**	0.003	13.5	ug/L	141	Standard
	Zn	66	529.7	2.7	**-0.3672**	0.009	2.6	ug/L	491	Standard
>	**Ge**	72	540586.6	1.6				ug/L	515087	Standard
	As	75	16.4	99.5	**0.0543**	0.014	25.8	ug/L	20	Standard
	Se	82	18.7	7.6	**0.0353**	0.014	40.6	ug/L	15	Standard
	Se-1	77	50.3	10.9	**-0.0819**	0.075	91.1	ug/L	48	Standard
>	**Ga**	71	86.7	18.5				mg/L	90	Standard
	Rb	85	13.3	78.1				ug/L	23	Standard
	Y	89	464568.8	2.3				ug/L	448510	Standard
>	**Rh**	103	18.3	15.7				ug/L	15	Standard
	Mo	98	439.4	21.7	**0.1377**	0.033	23.9	ug/L	11	Standard
	Ag	107	149.7	25.4	**0.0053**	0.005	102.3	ug/L	134	Standard
	Cd	111	15.2	39.6	**0.0040**	0.003	67.9	mg/L	7	Standard
	Cd	114	45.8	32.1	**0.0036**	0.003	74.9	ug/L	16	Standard
>	**In**	115	633054.7	0.7				ug/L	611117	Standard
	Sn	118	1025.0	34.7	**0.0655**	0.058	89.1	ug/L	518	Standard
	Sb	123	2042.3	25.6	**0.4053**	0.108	26.7	ug/L	127	Standard
	Ba	135	49.7	4.2	**-0.0019**	0.001	60.7	ug/L	45	Standard
	Ce	140	36.7	20.8				ug/L	298	Standard
>	**Tb**	159	1001668.5	0.7				ug/L	977942	Standard
	Ho	165	10.0					ug/L	12	Standard
	Tl	203	78.0	41.6	**0.0041**	0.004	88.8	ug/L	31	Standard
	Tl	205	58.3	47.2	**0.0054**	0.004	66.7	ug/L	20	Standard
	Pb	206	315.3	11.1	**-0.0075**	0.006	81.9	ug/L	286	Standard
	Pb	207	291.3	4.8	**-0.0005**	0.003	477.4	ug/L	257	Standard
	Pb	208	1079.0	5.9	**-0.0054**	0.003	55.4	ug/L	956	Standard
	U	238	63.3	42.9	**0.0055**	0.004	65.2	ug/L	17	Standard
>	**Bi**	209	578819.3	0.5				ug/L	557080	Standard

Sample ID: QC Std 2
Report Date/Time: Tuesday, June 28, 2016 13:24:27
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	25.0	40.0	**-0.0249**	0.013	50.9	mg/L	32	Standard
K	39	31.7	9.1	**-0.0375**	0.009	24.5	mg/L	30	Standard
Ca	43	41.7	36.7	**0.7901**	2.216	280.5	mg/L	48	Standard
Fe	54	146.0	8.6	**-0.0087**	0.010	113.7	mg/L	159	Standard
Fe	57	195.0	25.3	**0.0139**	0.156	1122.4	mg/L	195	Standard
Sc-1	45	40706.3	0.9				mg/L	39543	Standard
Cl	35	29554.8	2.8				ug/L	27642	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	1693.4	6.8				ug/L	1647	Standard
P	31	17536.8	2.2				ug/L	15231	Standard
S	34	2233.5	3.0				ug/L	1845	Standard
Sr	88	66.7	15.6				ug/L	70	Standard
C	12	96.7	33.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	12.7	49.8				mg/L	12	Standard
Ho-1	165	10.0					mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	1260.1	15.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		104.951	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 2
Report Date/Time: Tuesday, June 28, 2016 13:24:27
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.590
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.902
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 2	Sb	123	
QC Std 2	Ca	43	

Sample ID: QC Std 2
Report Date/Time: Tuesday, June 28, 2016 13:24:27
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 3
Sample Date/Time: Tuesday, June 28, 2016 13:25:23
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79486.3	1.5				ug/L	81760	Standard
	Be	9	236.7	23.6	**0.2350**	0.054	23.1	ug/L	17	Standard
	Al	27	1635.1	14.9	**-0.0035**	0.001	42.0	ug/L	1410	Standard
	Sc	45	39860.7	2.7				ug/L	39543	Standard
	Ti	47	39.7	2.9	**-0.0059**	0.005	91.5	ug/L	39	Standard
	V	51	3750.6	3.6	**0.3321**	0.019	5.7	ug/L	1415	Standard
	Cr	52	11750.6	2.2	**0.6283**	0.044	7.0	ug/L	6946	Standard
	Cr	53	1020.0	3.7	**0.8185**	0.038	4.7	ug/L	328	Standard
	Mn	55	3800.5	0.2	**0.4368**	0.004	0.8	ug/L	823	Standard
	Co	59	2925.3	4.6	**0.3843**	0.020	5.3	ug/L	252	Standard
	Ni	60	3080.3	1.6	**1.5280**	0.033	2.2	ug/L	192	Standard
	Cu	65	1617.4	2.7	**0.7591**	0.027	3.5	ug/L	141	Standard
	Zn	66	7652.9	2.4	**6.1434**	0.191	3.1	ug/L	491	Standard
>	**Ge**	72	515947.7	0.4				ug/L	515087	Standard
	As	75	426.1	10.9	**0.4186**	0.040	9.6	ug/L	20	Standard
	Se	82	65.5	20.2	**0.4487**	0.114	25.5	ug/L	15	Standard
	Se-1	77	77.0	13.5	**0.3017**	0.140	46.4	ug/L	48	Standard
>	**Ga**	71	75.0	29.1				mg/L	90	Standard
	Rb	85	11.7	49.5				ug/L	23	Standard
	Y	89	454544.4	2.9				ug/L	448510	Standard
>	**Rh**	103	11.7	65.5				ug/L	15	Standard
	Mo	98	144.3	6.8	**0.0328**	0.004	11.5	ug/L	11	Standard
	Ag	107	2677.9	4.8	**0.3715**	0.022	5.8	ug/L	134	Standard
	Cd	111	522.1	8.6	**0.2344**	0.020	8.5	mg/L	7	Standard
	Cd	114	1313.6	1.7	**0.2408**	0.006	2.4	ug/L	16	Standard
>	**In**	115	628989.1	0.7				ug/L	611117	Standard
	Sn	118	611.7	1.9	**-0.0014**	0.001	96.1	ug/L	518	Standard
	Sb	123	2074.6	8.0	**0.4143**	0.033	7.9	ug/L	127	Standard
	Ba	135	1500.7	4.3	**0.6813**	0.036	5.2	ug/L	45	Standard
	Ce	140	23.3	24.7				ug/L	298	Standard
>	**Tb**	159	1002704.6	1.0				ug/L	977942	Standard
	Ho	165	21.7	35.3				ug/L	12	Standard
	Tl	203	715.4	2.1	**0.0768**	0.002	2.1	ug/L	31	Standard
	Tl	205	653.3	9.4	**0.0835**	0.008	10.1	ug/L	20	Standard
	Pb	206	1358.4	2.4	**0.1778**	0.007	3.8	ug/L	286	Standard
	Pb	207	1172.0	3.3	**0.1729**	0.010	5.8	ug/L	257	Standard
	Pb	208	4784.3	0.9	**0.1832**	0.004	2.4	ug/L	956	Standard
	U	238	2840.3	2.6	**0.3778**	0.014	3.6	ug/L	17	Standard
>	**Bi**	209	580093.5	1.0				ug/L	557080	Standard

Sample ID: QC Std 3
Report Date/Time: Tuesday, June 28, 2016 13:27:40
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	33.3	22.9	**-0.0129**	0.011	85.9	mg/L	32	Standard
K	39	38.3	45.8	**-0.0130**	0.059	454.3	mg/L	30	Standard
Ca	43	38.3	41.9	**0.4748**	2.540	535.0	mg/L	48	Standard
Fe	54	103.9	16.3	**-0.0436**	0.016	37.7	mg/L	159	Standard
Fe	57	210.0	19.5	**0.0806**	0.150	185.6	mg/L	195	Standard
> **Sc-1**	45	39860.7	2.7				mg/L	39543	Standard
Cl	35	28271.0	1.2				ug/L	27642	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	1630.1	3.7				ug/L	1647	Standard
P	31	12470.1	4.9				ug/L	15231	Standard
S	34	2090.1	6.4				ug/L	1845	Standard
Sr	88	61.7	72.7				ug/L	70	Standard
C	12	146.7	7.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	26.0	116.4				mg/L	12	Standard
Ho-1	165	21.7	35.3				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	1031.7	25.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	117.475		
Al	27	-0.351		
Sc	45			
Ti	47			
V	51	83.021		
Cr	52	78.533		
Cr	53			
Mn	55	87.355		
Co	59	96.082		
Ni	60	95.499		
Cu	65	94.881		
Zn	66	98.294		
> Ge	72		100.167	
As	75	104.644		
Se	82	112.180		
Se-1	77			
> Ga	71			

Sample ID: QC Std 3
Report Date/Time: Tuesday, June 28, 2016 13:27:40
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	92.871		
Cd	111	97.662		
Cd	114			
> In	115			102.925
Sn	118			
Sb	123	103.570		
Ba	135	90.845		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	95.941		
Tl	205			
Pb	206			
Pb	207			
Pb	208	91.584		
U	238	94.451		
> Bi	209			104.131
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 3	Al	27	

Sample ID: QC Std 3
Report Date/Time: Tuesday, June 28, 2016 13:27:40
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Tuesday, June 28, 2016 13:28:34
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	85379.9	0.7				ug/L	81760	Standard
	Be	9	28.3	44.4	**0.0176**	0.012	68.3	ug/L	17	Standard
	Al	27	7582301.1	0.5	**47.0918**	0.449	1.0	ug/L	1410	Standard
	Sc	45	42481.1	0.8				ug/L	39543	Standard
	Ti	47	26924.2	1.2	**102.1971**	2.644	2.6	ug/L	39	Standard
	V	51	1177.2	11.7	**-0.0466**	0.017	37.5	ug/L	1415	Standard
	Cr	52	7440.5	1.6	**-0.0468**	0.011	24.4	ug/L	6946	Standard
	Cr	53	2938.6	8.1	**2.8263**	0.310	11.0	ug/L	328	Standard
	Mn	55	1484.1	0.5	**0.0680**	0.007	9.9	ug/L	823	Standard
	Co	59	438.3	7.6	**0.0276**	0.004	13.4	ug/L	252	Standard
	Ni	60	870.4	2.5	**0.3318**	0.004	1.2	ug/L	192	Standard
	Cu	65	647.0	5.3	**0.2309**	0.010	4.5	ug/L	141	Standard
	Zn	66	2006.5	1.8	**0.9027**	0.019	2.1	ug/L	491	Standard
>	Ge	72	545025.5	2.5				ug/L	515087	Standard
	As	75	-62.2	34.4	**-0.0118**	0.017	144.3	ug/L	20	Standard
	Se	82	23.6	6.2	**0.0746**	0.009	12.7	ug/L	15	Standard
	Se-1	77	231.7	5.9	**2.1902**	0.222	10.1	ug/L	48	Standard
>	Ga	71	171.7	17.1				mg/L	90	Standard
	Rb	85	1090.0	4.4				ug/L	23	Standard
	Y	89	452797.1	1.4				ug/L	448510	Standard
>	Rh	103	10.0	100.0				ug/L	15	Standard
	Mo	98	263118.9	1.1	**94.9055**	0.578	0.6	ug/L	11	Standard
	Ag	107	286.7	3.7	**0.0254**	0.002	6.6	ug/L	134	Standard
	Cd	111	-226.3	7.3	**-0.1062**	0.009	8.1	mg/L	7	Standard
	Cd	114	466.4	13.0	**0.0826**	0.011	12.8	ug/L	16	Standard
>	In	115	626700.1	1.0				ug/L	611117	Standard
	Sn	118	941.7	5.7	**0.0533**	0.007	13.8	ug/L	518	Standard
	Sb	123	466.3	29.2	**0.0836**	0.028	33.8	ug/L	127	Standard
	Ba	135	616.7	6.0	**0.2662**	0.017	6.2	ug/L	45	Standard
	Ce	140	233.3	17.2				ug/L	298	Standard
>	Tb	159	1019781.5	0.4				ug/L	977942	Standard
	Ho	165	28.3	27.0				ug/L	12	Standard
	Tl	203	270.7	14.6	**0.0257**	0.004	17.2	ug/L	31	Standard
	Tl	205	205.0	17.1	**0.0243**	0.005	18.8	ug/L	20	Standard
	Pb	206	568.7	4.5	**0.0362**	0.004	12.0	ug/L	286	Standard
	Pb	207	508.0	2.4	**0.0408**	0.002	5.7	ug/L	257	Standard
	Pb	208	1902.0	8.6	**0.0352**	0.008	23.2	ug/L	956	Standard
	U	238	71.7	54.2	**0.0065**	0.005	78.6	ug/L	17	Standard
>	Bi	209	587460.1	0.4				ug/L	557080	Standard

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 13:30:51
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**29.8288**	29.665	99.5	mg/L	0	Standard
Mg	24	10050.0	1.2	**12.3961**	0.163	1.3	mg/L	32	Standard
K	39	1535.1	9.7	**4.6887**	0.503	10.7	mg/L	30	Standard
Ca	43	211.7	14.4	**24.5987**	4.529	18.4	mg/L	48	Standard
Fe	54	14184.2	1.5	**11.8132**	0.271	2.3	mg/L	159	Standard
Fe	57	3883.8	1.7	**11.5795**	0.109	0.9	mg/L	195	Standard
Sc-1	45	42481.1	0.8				mg/L	39543	Standard
Cl	35	27442.8	1.7				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1816.8	6.7				ug/L	1647	Standard
P	31	16744.2	0.2				ug/L	15231	Standard
S	34	1956.8	7.4				ug/L	1845	Standard
Sr	88	48.3	53.1				ug/L	70	Standard
C	12	220.0	25.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	56.7	27.0				mg/L	0	Standard
Dy	164	24.9	47.6				mg/L	12	Standard
Ho-1	165	28.3	27.0				mg/L	12	Standard
Er	166	36.7	56.8				mg/L	20	Standard
I	127	1896.8	6.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27	0.942		
Sc	45			
Ti	47	102.197		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		105.812	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 13:30:51
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	94.905		
Ag	107			
Cd	111			
Cd	114			
In	115			102.550
Sn	118			
Sb	123			
Ba	135			
Ce	140			
Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
Bi	209			105.453
Na	23	238.631		
Mg	24	247.921		
K	39	93.774		
Ca	43	163.991		
Fe	54	94.506		
Fe	57	92.636		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Na	23	
QC Std 4	Mg	24	

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 13:30:51
Page 3

Approved: June 29, 2016

QC Std 4	Ca	43

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 13:30:51
Page 4

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Tuesday, June 28, 2016 13:31:45
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87031.0	2.2				ug/L	81760	Standard
	Be	9	104600.0	3.1	**98.9445**	2.636	2.7	ug/L	17	Standard
	Al	27	8235764.9	2.8	**50.1881**	1.494	3.0	ug/L	1410	Standard
	Sc	45	40814.9	1.3				ug/L	39543	Standard
	Ti	47	28557.9	2.6	**110.3128**	3.806	3.5	ug/L	39	Standard
	V	51	698774.9	0.6	**96.5097**	2.203	2.3	ug/L	1415	Standard
	Cr	52	706026.0	2.3	**96.5109**	2.799	2.9	ug/L	6946	Standard
	Cr	53	91691.9	3.7	**100.1429**	4.249	4.2	ug/L	328	Standard
	Mn	55	659133.1	0.8	**97.6603**	2.181	2.2	ug/L	823	Standard
	Co	59	714164.5	1.5	**97.7493**	2.387	2.4	ug/L	252	Standard
	Ni	60	188241.9	2.2	**96.0473**	2.835	3.0	ug/L	192	Standard
	Cu	65	190773.3	1.3	**96.4911**	2.052	2.1	ug/L	141	Standard
	Zn	66	115799.2	0.5	**100.8065**	2.067	2.1	ug/L	491	Standard
>	**Ge**	72	535562.3	1.7				ug/L	515087	Standard
	As	75	115513.0	0.8	**98.8727**	1.879	1.9	ug/L	20	Standard
	Se	82	11807.8	0.8	**98.5471**	2.365	2.4	ug/L	15	Standard
	Se-1	77	7973.8	0.6	**101.0979**	1.926	1.9	ug/L	48	Standard
>	**Ga**	71	195.0	2.6				mg/L	90	Standard
	Rb	85	900.0	5.1				ug/L	23	Standard
	Y	89	467512.1	1.1				ug/L	448510	Standard
>	**Rh**	103	41.7	61.6				ug/L	15	Standard
	Mo	98	278616.6	1.0	**98.3192**	1.671	1.7	ug/L	11	Standard
	Ag	107	629653.8	2.5	**89.4581**	1.609	1.8	ug/L	134	Standard
	Cd	111	214179.9	0.7	**95.5849**	1.082	1.1	mg/L	7	Standard
	Cd	114	513207.5	0.9	**94.2229**	0.578	0.6	ug/L	16	Standard
>	**In**	115	640649.3	1.3				ug/L	611117	Standard
	Sn	118	1085.0	2.4	**0.0732**	0.006	8.8	ug/L	518	Standard
	Sb	123	468645.5	2.2	**94.7292**	1.132	1.2	ug/L	127	Standard
	Ba	135	204538.3	1.7	**94.4814**	1.088	1.2	ug/L	45	Standard
	Ce	140	98.3	7.8				ug/L	298	Standard
>	**Tb**	159	1030592.0	2.2				ug/L	977942	Standard
	Ho	165	16.7	45.8				ug/L	12	Standard
	Tl	203	850272.7	1.9	**96.3847**	2.377	2.5	ug/L	31	Standard
	Tl	205	753805.3	1.5	**98.4067**	2.429	2.5	ug/L	20	Standard
	Pb	206	560444.8	0.3	**98.9799**	0.961	1.0	ug/L	286	Standard
	Pb	207	485626.8	0.4	**95.0861**	1.259	1.3	ug/L	257	Standard
	Pb	208	1940003.2	1.2	**98.1909**	2.155	2.2	ug/L	956	Standard
	U	238	731621.6	1.7	**97.5330**	2.488	2.6	ug/L	17	Standard
>	**Bi**	209	583301.0	1.0				ug/L	557080	Standard

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 13:34:02
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**30.9951**	30.718	99.1	mg/L	0	Standard
Mg	24	10642.1	4.7	**13.6704**	0.697	5.1	mg/L	32	Standard
K	39	1555.1	3.7	**4.9498**	0.207	4.2	mg/L	30	Standard
Ca	43	191.7	19.9	**22.8900**	5.801	25.3	mg/L	48	Standard
Fe	54	15478.2	1.2	**13.4350**	0.150	1.1	mg/L	159	Standard
Fe	57	4137.2	0.8	**12.9082**	0.151	1.2	mg/L	195	Standard
Sc-1	45	40814.9	1.3				mg/L	39543	Standard
Cl	35	27689.3	2.5				ug/L	27642	Standard
Kr	83	2.0	100.0				ug/L	2	Standard
Br	81	1876.8	10.9				ug/L	1647	Standard
P	31	15799.9	3.3				ug/L	15231	Standard
S	34	1943.5	5.2				ug/L	1845	Standard
Sr	88	65.0	27.7				ug/L	70	Standard
C	12	256.7	13.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	15.9	127.8				mg/L	12	Standard
Ho-1	165	16.7	45.8				mg/L	12	Standard
Er	166	16.7	69.3				mg/L	20	Standard
I	127	1231.7	9.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	98.944		
Al	27	1.004		
Sc	45			
Ti	47	110.313		
V	51	96.510		
Cr	52	96.511		
Cr	53			
Mn	55	97.660		
Co	59	97.749		
Ni	60	96.047		
Cu	65	96.491		
Zn	66	100.807		
Ge	72		103.975	
As	75	98.873		
Se	82	98.547		
Se-1	77			
Ga	71			

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 13:34:02
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	98.319		
Ag	107	89.458		
Cd	111	95.585		
Cd	114			
In	115		104.833	
Sn	118			
Sb	123	94.729		
Ba	135	94.481		
Ce	140			
Tb	159			
Ho	165			
Tl	203	96.385		
Tl	205			
Pb	206			
Pb	207			
Pb	208	98.191		
U	238	97.533		
Bi	209		104.707	
Na	23	247.961		
Mg	24	273.408		
K	39	98.997		
Ca	43	152.600		
Fe	54	107.480		
Fe	57	103.265		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Na	23	
QC Std 5	Mg	24	

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 13:34:02
Page 3

Approved: June 29, 2016

QC Std 5	Ca	43

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 13:34:02
Page 4

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 13:34:59
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85572.8	3.1				ug/L	81760	Standard
	Be	9	51348.6	3.0	**49.4131**	1.801	3.6	ug/L	17	Standard
	Al	27	7869928.8	0.7	**48.7904**	1.162	2.4	ug/L	1410	Standard
	Sc	45	41500.1	1.3				ug/L	39543	Standard
	Ti	47	25257.0	1.4	**98.1904**	2.232	2.3	ug/L	39	Standard
	V	51	359717.6	0.9	**49.9155**	1.252	2.5	ug/L	1415	Standard
	Cr	52	362752.3	1.0	**49.4079**	0.954	1.9	ug/L	6946	Standard
	Cr	53	45483.2	1.3	**49.8439**	1.589	3.2	ug/L	328	Standard
	Mn	55	333977.4	0.1	**49.7416**	1.012	2.0	ug/L	823	Standard
	Co	59	371364.3	0.2	**51.1559**	1.087	2.1	ug/L	252	Standard
	Ni	60	96391.0	0.8	**49.4567**	0.760	1.5	ug/L	192	Standard
	Cu	65	97494.7	0.8	**49.5943**	0.620	1.2	ug/L	141	Standard
	Zn	66	56948.5	0.8	**49.4853**	0.762	1.5	ug/L	491	Standard
>	**Ge**	72	532023.1	2.0				ug/L	515087	Standard
	As	75	57437.4	0.6	**49.5161**	1.180	2.4	ug/L	20	Standard
	Se	82	5946.0	0.9	**49.9013**	1.415	2.8	ug/L	15	Standard
	Se-1	77	4004.5	1.6	**50.7471**	0.199	0.4	ug/L	48	Standard
>	**Ga**	71	93.3	16.4				mg/L	90	Standard
	Rb	85	881.7	8.3				ug/L	23	Standard
	Y	89	456900.1	0.4				ug/L	448510	Standard
>	**Rh**	103	26.7	60.3				ug/L	15	Standard
	Mo	98	274645.7	0.6	**96.2670**	1.141	1.2	ug/L	11	Standard
	Ag	107	344865.6	0.2	**48.6675**	0.302	0.6	ug/L	134	Standard
	Cd	111	109918.0	0.6	**48.7237**	0.167	0.3	mg/L	7	Standard
	Cd	114	266820.9	0.9	**48.6598**	0.652	1.3	ug/L	16	Standard
>	**In**	115	644937.8	0.8				ug/L	611117	Standard
	Sn	118	304766.5	1.2	**48.7367**	0.172	0.4	ug/L	518	Standard
	Sb	123	245072.6	0.1	**49.2072**	0.383	0.8	ug/L	127	Standard
	Ba	135	106223.8	1.1	**48.7299**	0.483	1.0	ug/L	45	Standard
	Ce	140	76.7	21.0				ug/L	298	Standard
>	**Tb**	159	1040424.8	0.9				ug/L	977942	Standard
	Ho	165	11.7	89.2				ug/L	12	Standard
	Tl	203	436254.3	0.9	**48.8621**	1.048	2.1	ug/L	31	Standard
	Tl	205	382952.9	1.1	**49.3898**	0.698	1.4	ug/L	20	Standard
	Pb	206	278752.3	0.9	**48.6168**	1.163	2.4	ug/L	286	Standard
	Pb	207	252906.4	1.0	**48.9055**	1.215	2.5	ug/L	257	Standard
	Pb	208	993060.2	1.0	**49.6339**	1.159	2.3	ug/L	956	Standard
	U	238	369537.8	0.4	**48.6731**	0.911	1.9	ug/L	17	Standard
>	**Bi**	209	590375.8	1.5				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 13:37:15
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4017.2	5.2	**5.0405**	0.318	6.3	mg/L	32	Standard
K	39	1600.1	5.0	**5.0128**	0.329	6.6	mg/L	30	Standard
Ca	43	98.3	19.3	**8.9105**	2.944	33.0	mg/L	48	Standard
Fe	54	5965.0	1.0	**5.0068**	0.016	0.3	mg/L	159	Standard
Fe	57	1785.1	5.1	**5.1208**	0.370	7.2	mg/L	195	Standard
> **Sc-1**	45	41500.1	1.3				mg/L	39543	Standard
Cl	35	29160.0	1.9				ug/L	27642	Standard
Kr	83	3.3	45.8				ug/L	2	Standard
Br	81	1670.1	8.2				ug/L	1647	Standard
P	31	17116.3	2.7				ug/L	15231	Standard
S	34	2186.8	7.3				ug/L	1845	Standard
Sr	88	41.7	34.6				ug/L	70	Standard
C	12	130.0	42.8				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	23.3	24.7				mg/L	12	Standard
Ho-1	165	11.7	89.2				mg/L	12	Standard
Er	166	0.0					mg/L	20	Standard
I	127	1351.7	14.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	98.826		
Al	27	97.581		
Sc	45			
Ti	47	98.190		
V	51	99.831		
Cr	52	98.816		
Cr	53			
Mn	55	99.483		
Co	59	102.312		
Ni	60	98.913		
Cu	65	99.189		
Zn	66	98.971		
> Ge	72		103.288	
As	75	99.032		
Se	82	99.803		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 13:37:15
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	96.267		
Ag	107	97.335		
Cd	111	97.447		
Cd	114			
> In	115			105.534
Sn	118	97.473		
Sb	123	98.414		
Ba	135	97.460		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	97.724		
Tl	205			
Pb	206			
Pb	207			
Pb	208	99.268		
U	238	97.346		
> Bi	209			105.977
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 13:37:15
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 13:38:10
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82763.8	0.7				ug/L	81760	Standard
	Be	9	40.0	57.3	**0.0300**	0.023	75.9	ug/L	17	Standard
	Al	27	4257.4	63.2	**0.0129**	0.017	134.4	ug/L	1410	Standard
	Sc	45	40355.3	1.7				ug/L	39543	Standard
	Ti	47	44.3	23.5	**0.0085**	0.041	476.8	ug/L	39	Standard
	V	51	1283.7	14.1	**-0.0268**	0.025	92.2	ug/L	1415	Standard
	Cr	52	6366.3	1.6	**-0.1663**	0.013	7.9	ug/L	6946	Standard
	Cr	53	461.7	3.5	**0.1707**	0.014	8.3	ug/L	328	Standard
	Mn	55	939.4	15.6	**-0.0074**	0.022	294.6	ug/L	823	Standard
	Co	59	428.3	39.2	**0.0280**	0.023	82.9	ug/L	252	Standard
	Ni	60	218.7	25.5	**0.0081**	0.029	354.3	ug/L	192	Standard
	Cu	65	184.3	18.2	**0.0036**	0.017	469.2	ug/L	141	Standard
	Zn	66	515.3	2.3	**-0.3696**	0.014	3.7	ug/L	491	Standard
>	**Ge**	72	528838.2	0.8				ug/L	515087	Standard
	As	75	8.3	324.4	**0.0475**	0.023	49.1	ug/L	20	Standard
	Se	82	19.9	12.2	**0.0491**	0.021	43.7	ug/L	15	Standard
	Se-1	77	55.0	4.8	**-0.0079**	0.029	369.8	ug/L	48	Standard
>	**Ga**	71	56.7	39.8				mg/L	90	Standard
	Rb	85	10.0					ug/L	23	Standard
	Y	89	469796.8	1.6				ug/L	448510	Standard
>	**Rh**	103	20.0	25.0				ug/L	15	Standard
	Mo	98	452.3	25.3	**0.1439**	0.042	29.5	ug/L	11	Standard
	Ag	107	307.7	67.7	**0.0286**	0.031	108.2	ug/L	134	Standard
	Cd	111	45.2	95.8	**0.0178**	0.020	112.8	mg/L	7	Standard
	Cd	114	114.8	113.1	**0.0168**	0.025	147.3	ug/L	16	Standard
>	**In**	115	628024.0	1.4				ug/L	611117	Standard
	Sn	118	940.0	23.7	**0.0531**	0.039	73.1	ug/L	518	Standard
	Sb	123	1312.1	39.0	**0.2588**	0.110	42.3	ug/L	127	Standard
	Ba	135	79.3	75.1	**0.0125**	0.029	229.7	ug/L	45	Standard
	Ce	140	28.3	36.7				ug/L	298	Standard
>	**Tb**	159	1003248.8	1.6				ug/L	977942	Standard
	Ho	165	15.0	88.2				ug/L	12	Standard
	Tl	203	265.0	107.1	**0.0261**	0.033	127.7	ug/L	31	Standard
	Tl	205	163.3	84.5	**0.0196**	0.019	95.5	ug/L	20	Standard
	Pb	206	421.0	42.3	**0.0127**	0.033	261.4	ug/L	286	Standard
	Pb	207	362.0	44.3	**0.0147**	0.033	224.6	ug/L	257	Standard
	Pb	208	1358.7	30.8	**0.0100**	0.022	223.6	ug/L	956	Standard
	U	238	139.0	84.9	**0.0161**	0.016	101.8	ug/L	17	Standard
>	**Bi**	209	570467.7	1.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 13:40:27
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.3221**	17.870	173.1	mg/L	0	Standard
Mg	24	30.0	28.9	**-0.0179**	0.012	67.0	mg/L	32	Standard
K	39	31.7	24.1	**-0.0367**	0.024	65.4	mg/L	30	Standard
Ca	43	38.3	15.1	**0.3686**	0.952	258.2	mg/L	48	Standard
Fe	54	115.7	16.2	**-0.0344**	0.017	49.5	mg/L	159	Standard
Fe	57	198.3	8.1	**0.0320**	0.061	191.3	mg/L	195	Standard
> **Sc-1**	45	40355.3	1.7				mg/L	39543	Standard
Cl	35	29509.4	2.2				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	1783.4	4.3				ug/L	1647	Standard
P	31	17421.7	1.2				ug/L	15231	Standard
S	34	2170.2	3.8				ug/L	1845	Standard
Sr	88	55.0	15.7				ug/L	70	Standard
C	12	93.3	40.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	8.9	120.7				mg/L	12	Standard
Ho-1	165	15.0	88.2				mg/L	12	Standard
Er	166	23.3	65.5				mg/L	20	Standard
I	127	1233.4	14.2				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		102.670	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 13:40:27
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.767
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.403
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Sb	123	

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 13:40:27
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036203
Sample Date/Time: Tuesday, June 28, 2016 13:41:22
Number of Replicates: 3
Autosampler Position: 301
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83400.5	10.2				ug/L	81760	Standard
	Be	9	11.7	65.5	**0.0015**	0.006	425.0	ug/L	17	Standard
	Al	27	3017.0	39.5	**0.0053**	0.009	165.2	ug/L	1410	Standard
	Sc	45	40218.8	12.0				ug/L	39543	Standard
	Ti	47	37.0	23.6	**-0.0163**	0.022	135.1	ug/L	39	Standard
	V	51	613.8	4.6	**-0.1167**	0.012	10.1	ug/L	1415	Standard
	Cr	52	3481.1	12.6	**-0.5531**	0.019	3.4	ug/L	6946	Standard
	Cr	53	250.0	27.5	**-0.0570**	0.057	99.8	ug/L	328	Standard
	Mn	55	1041.4	26.2	**0.0117**	0.027	231.5	ug/L	823	Standard
	Co	59	254.7	5.9	**0.0055**	0.004	71.3	ug/L	252	Standard
	Ni	60	206.3	13.3	**0.0056**	0.009	153.1	ug/L	192	Standard
	Cu	65	131.7	19.5	**-0.0206**	0.013	65.4	ug/L	141	Standard
	Zn	66	531.7	22.4	**-0.3419**	0.069	20.1	ug/L	491	Standard
>	**Ge**	72	510441.4	10.5				ug/L	515087	Standard
	As	75	44.6	25.0	**0.0799**	0.006	7.8	ug/L	20	Standard
	Se	82	38.3	36.8	**0.2107**	0.093	44.1	ug/L	15	Standard
	Se-1	77	32.0	11.3	**-0.2872**	0.062	21.6	ug/L	48	Standard
>	**Ga**	71	61.7	4.7				mg/L	90	Standard
	Rb	85	33.3	37.7				ug/L	23	Standard
	Y	89	399549.3	20.8				ug/L	448510	Standard
>	**Rh**	103	11.7	65.5				ug/L	15	Standard
	Mo	98	155.9	22.2	**0.0482**	0.023	47.8	ug/L	11	Standard
	Ag	107	121.0	24.4	**0.0044**	0.005	118.2	ug/L	134	Standard
	Cd	111	13.4	45.5	**0.0046**	0.004	96.5	mg/L	7	Standard
	Cd	114	43.9	10.7	**0.0047**	0.001	14.6	ug/L	16	Standard
>	**In**	115	541793.5	15.9				ug/L	611117	Standard
	Sn	118	351.7	31.7	**-0.0361**	0.012	32.5	ug/L	518	Standard
	Sb	123	283.1	30.2	**0.0538**	0.012	21.7	ug/L	127	Standard
	Ba	135	37.0	32.8	**-0.0052**	0.004	82.4	ug/L	45	Standard
	Ce	140	30.0	28.9				ug/L	298	Standard
>	**Tb**	159	879683.0	16.9				ug/L	977942	Standard
	Ho	165	8.3	34.6				ug/L	12	Standard
	Tl	203	99.7	75.1	**0.0097**	0.013	137.9	ug/L	31	Standard
	Tl	205	85.0	86.7	**0.0121**	0.015	120.6	ug/L	20	Standard
	Pb	206	313.0	1.9	**0.0011**	0.012	1036.8	ug/L	286	Standard
	Pb	207	266.0	5.5	**0.0034**	0.015	442.4	ug/L	257	Standard
	Pb	208	1233.7	12.3	**0.0133**	0.020	150.4	ug/L	956	Standard
	U	238	39.3	36.7	**0.0031**	0.002	60.3	ug/L	17	Standard
>	**Bi**	209	508733.8	16.9				ug/L	557080	Standard

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:43:39
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	21.7	13.3	**-0.0282**	0.008	27.2	mg/L	32	Standard
K	39	25.0	34.6	**-0.0591**	0.022	36.5	mg/L	30	Standard
Ca	43	40.0	12.5	**0.7483**	1.524	203.6	mg/L	48	Standard
Fe	54	113.2	25.1	**-0.0334**	0.039	117.8	mg/L	159	Standard
Fe	57	238.3	14.7	**0.1717**	0.127	73.7	mg/L	195	Standard
> **Sc-1**	45	40218.8	12.0				mg/L	39543	Standard
Cl	35	13831.7	10.0				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	810.0	8.9				ug/L	1647	Standard
P	31	10527.1	24.9				ug/L	15231	Standard
S	34	970.0	6.2				ug/L	1845	Standard
Sr	88	66.7	11.5				ug/L	70	Standard
C	12	110.0	9.1				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	12.4	124.8				mg/L	12	Standard
Ho-1	165	8.3	34.6				mg/L	12	Standard
Er	166	20.0	50.0				mg/L	20	Standard
I	127	523.3	19.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.006	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.098	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:43:39
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		88.656
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		91.321
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Cr 52 Lower	Cr	52	

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:43:39
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036204
Sample Date/Time: Tuesday, June 28, 2016 13:44:34
Number of Replicates: 3
Autosampler Position: 302
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83225.5	7.5				ug/L	81760	Standard
	Be	9	31228.6	61.6	**31.9371**	21.971	68.8	ug/L	17	Standard
	Al	27	8471.1	48.6	**0.0411**	0.031	75.9	ug/L	1410	Standard
	Sc	45	40830.0	2.2				ug/L	39543	Standard
	Ti	47	34.3	11.8	**-0.0297**	0.018	59.1	ug/L	39	Standard
	V	51	224078.9	58.8	**31.3360**	18.302	58.4	ug/L	1415	Standard
	Cr	52	238015.8	57.7	**32.3999**	19.040	58.8	ug/L	6946	Standard
	Cr	53	29690.7	58.7	**32.7502**	19.168	58.5	ug/L	328	Standard
	Mn	55	213724.7	56.8	**32.1044**	18.055	56.2	ug/L	823	Standard
	Co	59	222487.2	58.7	**30.9454**	17.938	58.0	ug/L	252	Standard
	Ni	60	62090.3	57.3	**32.1492**	18.240	56.7	ug/L	192	Standard
	Cu	65	63757.0	57.0	**32.7355**	18.436	56.3	ug/L	141	Standard
	Zn	66	38227.9	57.9	**33.2873**	19.476	58.5	ug/L	491	Standard
>	**Ge**	72	525473.2	1.2				ug/L	515087	Standard
	As	75	36453.5	56.6	**31.7527**	17.695	55.7	ug/L	20	Standard
	Se	82	3780.3	55.0	**31.9968**	17.401	54.4	ug/L	15	Standard
	Se-1	77	2538.3	56.6	**32.2335**	18.366	57.0	ug/L	48	Standard
>	**Ga**	71	76.7	26.4				mg/L	90	Standard
	Rb	85	43.3	17.6				ug/L	23	Standard
	Y	89	474726.4	7.6				ug/L	448510	Standard
>	**Rh**	103	26.7	21.7				ug/L	15	Standard
	Mo	98	564.0	40.4	**0.1823**	0.089	49.0	ug/L	11	Standard
	Ag	107	212096.4	59.4	**30.6350**	19.274	62.9	ug/L	134	Standard
	Cd	111	68030.3	54.2	**30.8198**	17.865	58.0	mg/L	7	Standard
	Cd	114	165078.2	56.3	**30.7839**	18.487	60.1	ug/L	16	Standard
>	**In**	115	641067.2	4.2				ug/L	611117	Standard
	Sn	118	550.0	28.6	**-0.0125**	0.029	234.3	ug/L	518	Standard
	Sb	123	143570.7	60.9	**29.5172**	18.974	64.3	ug/L	127	Standard
	Ba	135	66133.4	54.2	**31.0035**	17.964	57.9	ug/L	45	Standard
	Ce	140	83.3	81.7				ug/L	298	Standard
>	**Tb**	159	1026938.7	4.8				ug/L	977942	Standard
	Ho	165	11.7	65.5				ug/L	12	Standard
	Tl	203	274073.1	56.4	**30.7538**	17.663	57.4	ug/L	31	Standard
	Tl	205	243234.0	54.4	**31.4411**	17.425	55.4	ug/L	20	Standard
	Pb	206	180330.9	55.5	**31.4753**	17.798	56.5	ug/L	286	Standard
	Pb	207	158009.7	55.1	**30.5868**	17.206	56.3	ug/L	257	Standard
	Pb	208	646417.5	54.5	**32.3514**	17.977	55.6	ug/L	956	Standard
	U	238	234720.5	55.2	**30.9733**	17.414	56.2	ug/L	17	Standard
>	**Bi**	209	588853.8	4.4				ug/L	557080	Standard

Sample ID: L1606036204
Report Date/Time: Tuesday, June 28, 2016 13:46:51
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.1717**	17.609	173.1	mg/L	0	Standard
Mg	24	26.7	10.8	**-0.0228**	0.004	16.0	mg/L	32	Standard
K	39	36.7	43.8	**-0.0214**	0.052	244.2	mg/L	30	Standard
Ca	43	30.0	66.7	**-0.9172**	2.942	320.7	mg/L	48	Standard
Fe	54	107.6	36.8	**-0.0427**	0.035	81.3	mg/L	159	Standard
Fe	57	220.0	9.9	**0.0952**	0.078	81.4	mg/L	195	Standard
> **Sc-1**	45	40830.0	2.2				mg/L	39543	Standard
Cl	35	20640.2	29.5				ug/L	27642	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	1230.1	37.8				ug/L	1647	Standard
P	31	11926.4	10.7				ug/L	15231	Standard
S	34	1536.7	28.4				ug/L	1845	Standard
Sr	88	66.7	11.5				ug/L	70	Standard
C	12	146.7	27.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	5.9	100.8				mg/L	12	Standard
Ho-1	165	11.7	65.5				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	785.0	24.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.792	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.016	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036204
Report Date/Time: Tuesday, June 28, 2016 13:46:51
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.901
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.704
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036203
Sample Date/Time: Tuesday, June 28, 2016 13:49:03
Number of Replicates: 3
Autosampler Position: 205
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90942.6	5.2				ug/L	81760	Standard
	Be	9	53.3	85.1	**0.0399**	0.045	111.9	ug/L	17	Standard
	Al	27	693137.8	2.1	**4.0335**	0.159	3.9	ug/L	1410	Standard
	Sc	45	39909.2	1.1				ug/L	39543	Standard
	Ti	47	362.3	4.5	**1.2955**	0.068	5.3	ug/L	39	Standard
	V	51	2278.9	9.8	**0.1220**	0.031	25.8	ug/L	1415	Standard
	Cr	52	6909.6	1.4	**-0.0622**	0.012	19.7	ug/L	6946	Standard
	Cr	53	1090.0	12.5	**0.9032**	0.155	17.1	ug/L	328	Standard
	Mn	55	5526773.3	1.7	**854.3657**	14.322	1.7	ug/L	823	Standard
	Co	59	9370.2	5.6	**1.3053**	0.073	5.6	ug/L	252	Standard
	Ni	60	2122.5	1.2	**1.0247**	0.015	1.4	ug/L	192	Standard
	Cu	65	2498.9	0.1	**1.2274**	0.002	0.2	ug/L	141	Standard
	Zn	66	6516.4	0.8	**5.1316**	0.039	0.8	ug/L	491	Standard
>	**Ge**	72	513880.5	0.2				ug/L	515087	Standard
	As	75	3228.7	3.4	**2.9189**	0.105	3.6	ug/L	20	Standard
	Se	82	922.2	2.2	**7.9103**	0.196	2.5	ug/L	15	Standard
	Se-1	77	155.0	10.7	**1.3431**	0.217	16.2	ug/L	48	Standard
>	**Ga**	71	388.3	17.5				mg/L	90	Standard
	Rb	85	68372.4	3.2				ug/L	23	Standard
	Y	89	447009.0	1.5				ug/L	448510	Standard
>	**Rh**	103	111.7	20.2				ug/L	15	Standard
	Mo	98	1200247.1	0.8	**451.3790**	8.919	2.0	ug/L	11	Standard
	Ag	107	153.3	18.9	**0.0070**	0.004	61.8	ug/L	134	Standard
	Cd	111	-1054.6	0.6	**-0.5044**	0.004	0.7	mg/L	7	Standard
	Cd	114	1800.6	2.5	**0.3474**	0.013	3.8	ug/L	16	Standard
>	**In**	115	601262.2	1.3				ug/L	611117	Standard
	Sn	118	788.4	4.6	**0.0336**	0.006	17.8	ug/L	518	Standard
	Sb	123	1882.7	34.8	**0.3916**	0.136	34.8	ug/L	127	Standard
	Ba	135	4520.7	1.0	**2.2009**	0.051	2.3	ug/L	45	Standard
	Ce	140	6458.0	2.1				ug/L	298	Standard
>	**Tb**	159	1000527.6	0.4				ug/L	977942	Standard
	Ho	165	130.0	21.4				ug/L	12	Standard
	Tl	203	285.7	2.5	**0.0299**	0.000	1.6	ug/L	31	Standard
	Tl	205	403.3	76.1	**0.0543**	0.043	79.9	ug/L	20	Standard
	Pb	206	899.7	2.3	**0.1068**	0.005	4.7	ug/L	286	Standard
	Pb	207	732.0	4.6	**0.0957**	0.009	9.0	ug/L	257	Standard
	Pb	208	3524.8	26.8	**0.1310**	0.052	40.0	ug/L	956	Standard
	U	238	352.3	76.1	**0.0475**	0.039	81.4	ug/L	17	Standard
>	**Bi**	209	544253.0	1.1				ug/L	557080	Standard

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:51:20
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.4757**	18.136	173.1	mg/L	0	Standard
Mg	24	122834.3	0.9	**161.9583**	0.885	0.5	mg/L	32	Standard
K	39	4865.8	5.3	**16.1536**	1.001	6.2	mg/L	30	Standard
Ca	43	186.7	17.4	**22.7843**	5.205	22.8	mg/L	48	Standard
Fe	54	927.7	8.4	**0.6952**	0.076	10.9	mg/L	159	Standard
Fe	57	461.7	8.7	**0.9208**	0.151	16.4	mg/L	195	Standard
> **Sc-1**	45	39909.2	1.1				mg/L	39543	Standard
Cl	35	47371.9	1.0				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	196320.5	1.2				ug/L	1647	Standard
P	31	16525.7	2.7				ug/L	15231	Standard
S	34	2001.8	3.8				ug/L	1845	Standard
Sr	88	358.3	4.3				ug/L	70	Standard
C	12	246.7	19.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	211.1	26.7				mg/L	12	Standard
Ho-1	165	130.0	21.4				mg/L	12	Standard
Er	166	116.7	51.7				mg/L	20	Standard
I	127	1244078.5	12.9				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		111.231	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		99.766	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:51:20
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.387
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	97.697
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Mo 98 Upper, S, EEE	Mo	98	
Cd 111 Lower	Cd	111	

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:51:20
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036204
Sample Date/Time: Tuesday, June 28, 2016 13:52:18
Number of Replicates: 3
Autosampler Position: 206
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	103725.5	0.9				ug/L	81760	Standard
	Be	9	18.3	41.7	**0.0048**	0.006	125.1	ug/L	17	Standard
	Al	27	23803073.4	2.1	**121.7007**	1.863	1.5	ug/L	1410	Standard
	Sc	45	41989.8	2.5				ug/L	39543	Standard
	Ti	47	198.3	8.2	**0.6152**	0.056	9.1	ug/L	39	Standard
	V	51	2022.8	10.2	**0.0784**	0.030	38.8	ug/L	1415	Standard
	Cr	52	8025.5	3.2	**0.0718**	0.049	68.1	ug/L	6946	Standard
	Cr	53	1725.1	14.9	**1.5845**	0.307	19.4	ug/L	328	Standard
	Mn	55	8347.3	4.1	**1.1117**	0.038	3.4	ug/L	823	Standard
	Co	59	1991.1	1.1	**0.2461**	0.003	1.1	ug/L	252	Standard
	Ni	60	2784.9	2.6	**1.3428**	0.049	3.7	ug/L	192	Standard
	Cu	65	1676.4	9.2	**0.7724**	0.072	9.4	ug/L	141	Standard
	Zn	66	3951.5	5.5	**2.7003**	0.153	5.7	ug/L	491	Standard
>	**Ge**	72	526193.8	1.1				ug/L	515087	Standard
	As	75	658.8	2.1	**0.6140**	0.018	3.0	ug/L	20	Standard
	Se	82	229.7	4.0	**1.8340**	0.092	5.0	ug/L	15	Standard
	Se-1	77	153.0	9.2	**1.2707**	0.204	16.0	ug/L	48	Standard
>	**Ga**	71	148.3	16.6				mg/L	90	Standard
	Rb	85	683.3	4.9				ug/L	23	Standard
	Y	89	453248.0	2.1				ug/L	448510	Standard
>	**Rh**	103	91.7	27.5				ug/L	15	Standard
	Mo	98	5026.0	6.3	**1.8280**	0.098	5.4	ug/L	11	Standard
	Ag	107	95.3	7.7	**-0.0021**	0.001	57.5	ug/L	134	Standard
	Cd	111	8.0	39.8	**0.0008**	0.001	178.0	mg/L	7	Standard
	Cd	114	50.5	30.0	**0.0047**	0.003	60.8	ug/L	16	Standard
>	**In**	115	615030.5	1.6				ug/L	611117	Standard
	Sn	118	1100.0	6.8	**0.0830**	0.014	16.4	ug/L	518	Standard
	Sb	123	686.5	31.6	**0.1319**	0.046	34.9	ug/L	127	Standard
	Ba	135	22769.8	0.9	**10.9347**	0.080	0.7	ug/L	45	Standard
	Ce	140	320.0	14.9				ug/L	298	Standard
>	**Tb**	159	1016190.3	1.0				ug/L	977942	Standard
	Ho	165	48.3	15.8				ug/L	12	Standard
	Tl	203	156.0	5.7	**0.0132**	0.001	7.6	ug/L	31	Standard
	Tl	205	153.3	14.7	**0.0181**	0.003	17.2	ug/L	20	Standard
	Pb	206	466.3	6.0	**0.0202**	0.005	24.4	ug/L	286	Standard
	Pb	207	405.3	5.2	**0.0227**	0.004	15.6	ug/L	257	Standard
	Pb	208	1478.0	2.2	**0.0157**	0.002	13.7	ug/L	956	Standard
	U	238	1957.8	2.0	**0.2625**	0.004	1.4	ug/L	17	Standard
>	**Bi**	209	573383.4	0.9				ug/L	557080	Standard

Sample ID: L1606036204
Report Date/Time: Tuesday, June 28, 2016 13:54:35
Page 1

Approved: June 29, 2016

Na	23	8.3	34.6	**50.1179**	16.456	32.8	mg/L	0	Standard
Mg	24	58420.8	1.2	**73.2244**	2.756	3.8	mg/L	32	Standard
K	39	723.4	8.7	**2.1601**	0.187	8.7	mg/L	30	Standard
Ca	43	463.3	11.9	**60.9213**	7.520	12.3	mg/L	48	Standard
Fe	54	191.2	7.7	**0.0260**	0.014	55.5	mg/L	159	Standard
Fe	57	531.7	2.7	**1.0656**	0.010	0.9	mg/L	195	Standard
> **Sc-1**	45	41989.8	2.5				mg/L	39543	Standard
Cl	35	37603.3	1.8				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	30841.4	4.4				ug/L	1647	Standard
P	31	21757.3	1.3				ug/L	15231	Standard
S	34	1778.4	1.5				ug/L	1845	Standard
Sr	88	246.7	9.1				ug/L	70	Standard
C	12	216.7	7.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	60.8	24.7				mg/L	12	Standard
Ho-1	165	48.3	15.8				mg/L	12	Standard
Er	166	53.3	10.8				mg/L	20	Standard
I	127	78536.8	23.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		126.866	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.156	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036204
Report Date/Time: Tuesday, June 28, 2016 13:54:35
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	100.640
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.927
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606036204
Report Date/Time: Tuesday, June 28, 2016 13:54:35
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036205
Sample Date/Time: Tuesday, June 28, 2016 13:59:25
Number of Replicates: 3
Autosampler Position: 207
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	103215.4	0.9				ug/L	81760	Standard
	Be	9	20.0	50.0	**0.0063**	0.008	129.0	ug/L	17	Standard
	Al	27	14439258.9	3.3	**74.2049**	3.024	4.1	ug/L	1410	Standard
	Sc	45	39693.6	2.3				ug/L	39543	Standard
	Ti	47	140.0	4.0	**0.4180**	0.020	4.8	ug/L	39	Standard
	V	51	3604.7	5.2	**0.3305**	0.025	7.7	ug/L	1415	Standard
	Cr	52	7555.2	2.8	**0.0671**	0.025	36.8	ug/L	6946	Standard
	Cr	53	1441.7	2.8	**1.3602**	0.060	4.4	ug/L	328	Standard
	Mn	55	692335.2	1.4	**110.4896**	2.247	2.0	ug/L	823	Standard
	Co	59	1760.4	3.6	**0.2283**	0.011	4.9	ug/L	252	Standard
	Ni	60	2996.0	4.5	**1.5433**	0.085	5.5	ug/L	192	Standard
	Cu	65	1684.1	4.6	**0.8273**	0.040	4.8	ug/L	141	Standard
	Zn	66	2371.2	1.2	**1.4134**	0.020	1.4	ug/L	491	Standard
>	**Ge**	72	497232.9	0.8				ug/L	515087	Standard
	As	75	475.6	6.3	**0.4785**	0.029	6.0	ug/L	20	Standard
	Se	82	111.0	8.1	**0.8794**	0.084	9.6	ug/L	15	Standard
	Se-1	77	118.0	5.1	**0.9042**	0.095	10.5	ug/L	48	Standard
>	**Ga**	71	108.3	35.3				mg/L	90	Standard
	Rb	85	57350.0	0.9				ug/L	23	Standard
	Y	89	457066.8	2.1				ug/L	448510	Standard
>	**Rh**	103	28.3	44.4				ug/L	15	Standard
	Mo	98	255967.1	1.2	**93.6404**	2.667	2.8	ug/L	11	Standard
	Ag	107	89.3	9.3	**-0.0031**	0.001	44.5	ug/L	134	Standard
	Cd	111	-197.3	3.0	**-0.0942**	0.001	1.4	mg/L	7	Standard
	Cd	114	425.0	6.1	**0.0760**	0.006	8.0	ug/L	16	Standard
>	**In**	115	618093.1	1.7				ug/L	611117	Standard
	Sn	118	871.7	7.6	**0.0437**	0.009	20.3	ug/L	518	Standard
	Sb	123	474.5	27.0	**0.0866**	0.026	30.6	ug/L	127	Standard
	Ba	135	13314.9	2.5	**6.3544**	0.263	4.1	ug/L	45	Standard
	Ce	140	228.3	6.3				ug/L	298	Standard
>	**Tb**	159	1010942.5	1.3				ug/L	977942	Standard
	Ho	165	25.0	72.1				ug/L	12	Standard
	Tl	203	239.7	6.5	**0.0241**	0.002	9.5	ug/L	31	Standard
	Tl	205	181.7	22.1	**0.0230**	0.006	26.1	ug/L	20	Standard
	Pb	206	319.7	6.3	**-0.0035**	0.004	108.4	ug/L	286	Standard
	Pb	207	282.7	2.0	**0.0009**	0.002	252.3	ug/L	257	Standard
	Pb	208	1120.3	3.0	**-0.0001**	0.003	3300.1	ug/L	956	Standard
	U	238	733.7	3.0	**0.1011**	0.005	4.9	ug/L	17	Standard
>	**Bi**	209	548318.3	1.8				ug/L	557080	Standard

Sample ID: L1606036205
Report Date/Time: Tuesday, June 28, 2016 14:01:41
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**21.6860**	37.553	173.2	mg/L	0	Standard
Mg	24	97569.9	1.9	**129.3694**	3.560	2.8	mg/L	32	Standard
K	39	4389.0	6.2	**14.6373**	1.001	6.8	mg/L	30	Standard
Ca	43	378.3	2.8	**51.9377**	2.720	5.2	mg/L	48	Standard
Fe	54	155.8	7.4	**0.0035**	0.012	336.9	mg/L	159	Standard
Fe	57	446.7	11.3	**0.8756**	0.135	15.4	mg/L	195	Standard
> **Sc-1**	45	39693.6	2.3				mg/L	39543	Standard
Cl	35	45044.9	1.9				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	18554.7	4.8				ug/L	1647	Standard
P	31	20283.6	3.2				ug/L	15231	Standard
S	34	1950.1	2.9				ug/L	1845	Standard
Sr	88	140.0	14.3				ug/L	70	Standard
C	12	253.3	26.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	47.4	35.9				mg/L	12	Standard
Ho-1	165	25.0	72.1				mg/L	12	Standard
Er	166	53.3	21.7				mg/L	20	Standard
I	127	81681.7	5.6				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		126.242	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		96.534	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606036205
Report Date/Time: Tuesday, June 28, 2016 14:01:41
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	101.142	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	98.427	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036206
Sample Date/Time: Tuesday, June 28, 2016 14:02:36
Number of Replicates: 3
Autosampler Position: 208
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	108188.3	1.9				ug/L	81760	Standard
	Be	9	25.0	69.3	**0.0094**	0.014	144.0	ug/L	17	Standard
	Al	27	23235387.8	1.2	**113.9176**	1.672	1.5	ug/L	1410	Standard
	Sc	45	42060.0	3.5				ug/L	39543	Standard
	Ti	47	152.7	7.7	**0.4364**	0.055	12.6	ug/L	39	Standard
	V	51	2674.6	4.0	**0.1700**	0.012	6.8	ug/L	1415	Standard
	Cr	52	8527.1	1.1	**0.1424**	0.031	21.5	ug/L	6946	Standard
	Cr	53	1513.4	3.0	**1.3470**	0.073	5.5	ug/L	328	Standard
	Mn	55	17845.5	5.4	**2.5461**	0.135	5.3	ug/L	823	Standard
	Co	59	1119.4	15.5	**0.1245**	0.023	18.8	ug/L	252	Standard
	Ni	60	2702.9	1.5	**1.3004**	0.041	3.1	ug/L	192	Standard
	Cu	65	753.4	4.7	**0.2972**	0.013	4.4	ug/L	141	Standard
	Zn	66	1481.1	2.0	**0.4954**	0.035	7.0	ug/L	491	Standard
>	**Ge**	72	526170.8	1.4				ug/L	515087	Standard
	As	75	291.6	5.2	**0.2943**	0.014	4.7	ug/L	20	Standard
	Se	82	87.6	19.6	**0.6269**	0.155	24.7	ug/L	15	Standard
	Se-1	77	124.0	13.8	**0.8941**	0.239	26.7	ug/L	48	Standard
>	**Ga**	71	93.3	13.5				mg/L	90	Standard
	Rb	85	1890.1	1.5				ug/L	23	Standard
	Y	89	457488.4	1.5				ug/L	448510	Standard
>	**Rh**	103	41.7	45.4				ug/L	15	Standard
	Mo	98	1289.7	27.8	**0.4294**	0.125	29.1	ug/L	11	Standard
	Ag	107	132.0	62.6	**0.0022**	0.011	517.7	ug/L	134	Standard
	Cd	111	14.8	9.0	**0.0036**	0.001	18.2	mg/L	7	Standard
	Cd	114	35.3	64.8	**0.0015**	0.004	279.2	ug/L	16	Standard
>	**In**	115	650440.2	1.2				ug/L	611117	Standard
	Sn	118	795.0	5.8	**0.0244**	0.006	26.1	ug/L	518	Standard
	Sb	123	299.9	40.5	**0.0467**	0.023	50.3	ug/L	127	Standard
	Ba	135	16583.1	1.2	**7.5223**	0.109	1.5	ug/L	45	Standard
	Ce	140	573.3	12.7				ug/L	298	Standard
>	**Tb**	159	1070413.0	0.5				ug/L	977942	Standard
	Ho	165	35.0	14.3				ug/L	12	Standard
	Tl	203	143.0	68.4	**0.0109**	0.011	100.1	ug/L	31	Standard
	Tl	205	173.3	55.8	**0.0194**	0.012	61.7	ug/L	20	Standard
	Pb	206	407.0	5.8	**0.0055**	0.005	85.2	ug/L	286	Standard
	Pb	207	361.0	20.9	**0.0099**	0.014	143.4	ug/L	257	Standard
	Pb	208	1551.4	24.3	**0.0150**	0.018	119.8	ug/L	956	Standard
	U	238	2137.8	4.4	**0.2708**	0.011	4.2	ug/L	17	Standard
>	**Bi**	209	607194.4	1.0				ug/L	557080	Standard

Sample ID: L1606036206
Report Date/Time: Tuesday, June 28, 2016 14:05:08
Page 1

Approved: June 29, 2016

Na	23	15.0	57.7	**91.7381**	54.279	59.2	mg/L	0	Standard
Mg	24	7570.2	0.6	**9.4255**	0.372	3.9	mg/L	32	Standard
K	39	228.3	12.1	**0.5833**	0.074	12.7	mg/L	30	Standard
Ca	43	403.3	9.3	**52.3767**	6.757	12.9	mg/L	48	Standard
Fe	54	189.4	6.8	**0.0240**	0.007	29.2	mg/L	159	Standard
Fe	57	486.7	15.4	**0.9208**	0.241	26.2	mg/L	195	Standard
> **Sc-1**	45	42060.0	3.5				mg/L	39543	Standard
Cl	35	35390.6	1.2				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	8392.3	5.9				ug/L	1647	Standard
P	31	21440.2	3.3				ug/L	15231	Standard
S	34	2026.8	8.3				ug/L	1845	Standard
Sr	88	143.3	29.0				ug/L	70	Standard
C	12	223.3	9.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	47.9	21.4				mg/L	12	Standard
Ho-1	165	35.0	14.3				mg/L	12	Standard
Er	166	43.3	35.3				mg/L	20	Standard
I	127	10155.1	10.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.324	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.152	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036206
Report Date/Time: Tuesday, June 28, 2016 14:05:08
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.435
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.996
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036207
Sample Date/Time: Tuesday, June 28, 2016 14:06:02
Number of Replicates: 3
Autosampler Position: 209
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	107639.4	1.9				ug/L	81760	Standard
	Be	9	20.0	0.0	**0.0056**	0.000	5.4	ug/L	17	Standard
	Al	27	3476504.2	1.7	**17.1173**	0.043	0.3	ug/L	1410	Standard
	Sc	45	39367.8	2.7				ug/L	39543	Standard
	Ti	47	100.3	3.5	**0.2518**	0.007	3.0	ug/L	39	Standard
	V	51	5441.1	3.8	**0.6024**	0.020	3.4	ug/L	1415	Standard
	Cr	52	7747.6	3.6	**0.0931**	0.026	27.8	ug/L	6946	Standard
	Cr	53	1961.8	9.7	**1.9650**	0.153	7.8	ug/L	328	Standard
	Mn	55	12138.9	1.5	**1.7871**	0.034	1.9	ug/L	823	Standard
	Co	59	2054.5	1.6	**0.2711**	0.013	4.9	ug/L	252	Standard
	Ni	60	1564.4	4.1	**0.7535**	0.028	3.7	ug/L	192	Standard
	Cu	65	3547.1	3.0	**1.8404**	0.110	6.0	ug/L	141	Standard
	Zn	66	3174.7	2.0	**2.1684**	0.146	6.7	ug/L	491	Standard
>	**Ge**	72	498440.9	3.0				ug/L	515087	Standard
	As	75	2063.4	4.9	**1.9397**	0.147	7.6	ug/L	20	Standard
	Se	82	279.9	8.8	**2.3986**	0.288	12.0	ug/L	15	Standard
	Se-1	77	196.7	12.4	**1.9867**	0.408	20.5	ug/L	48	Standard
>	**Ga**	71	203.3	22.3				mg/L	90	Standard
	Rb	85	4590.7	1.6				ug/L	23	Standard
	Y	89	429873.2	1.1				ug/L	448510	Standard
>	**Rh**	103	36.7	15.7				ug/L	15	Standard
	Mo	98	411573.4	1.6	**157.7590**	2.905	1.8	ug/L	11	Standard
	Ag	107	97.0	12.7	**-0.0013**	0.002	144.7	ug/L	134	Standard
	Cd	111	-313.3	8.5	**-0.1547**	0.012	7.7	mg/L	7	Standard
	Cd	114	654.0	5.6	**0.1255**	0.007	5.6	ug/L	16	Standard
>	**In**	115	589803.8	0.9				ug/L	611117	Standard
	Sn	118	783.4	15.3	**0.0353**	0.020	57.5	ug/L	518	Standard
	Sb	123	727.4	9.3	**0.1468**	0.014	9.3	ug/L	127	Standard
	Ba	135	15565.3	1.0	**7.7872**	0.095	1.2	ug/L	45	Standard
	Ce	140	646.7	2.7				ug/L	298	Standard
>	**Tb**	159	991896.0	2.4				ug/L	977942	Standard
	Ho	165	86.7	18.5				ug/L	12	Standard
	Tl	203	276.7	3.3	**0.0282**	0.002	5.4	ug/L	31	Standard
	Tl	205	283.3	3.7	**0.0365**	0.002	4.5	ug/L	20	Standard
	Pb	206	399.0	5.4	**0.0104**	0.005	44.3	ug/L	286	Standard
	Pb	207	325.3	5.9	**0.0089**	0.004	45.8	ug/L	257	Standard
	Pb	208	1296.0	0.3	**0.0084**	0.001	12.8	ug/L	956	Standard
	U	238	1382.7	4.5	**0.1905**	0.008	4.1	ug/L	17	Standard
>	**Bi**	209	555677.7	1.4				ug/L	557080	Standard

Sample ID: L1606036207
Report Date/Time: Tuesday, June 28, 2016 14:08:19
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**21.6985**	18.802	86.6	mg/L	0	Standard
Mg	24	201508.1	2.0	**269.4503**	5.984	2.2	mg/L	32	Standard
K	39	1843.4	2.4	**6.1200**	0.302	4.9	mg/L	30	Standard
Ca	43	86.7	16.7	**7.9004**	2.376	30.1	mg/L	48	Standard
Fe	54	134.1	6.8	**-0.0152**	0.007	45.6	mg/L	159	Standard
Fe	57	218.3	31.3	**0.1119**	0.215	192.0	mg/L	195	Standard
> **Sc-1**	45	39367.8	2.7				mg/L	39543	Standard
Cl	35	57177.0	0.5				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	35885.9	6.7				ug/L	1647	Standard
P	31	17845.5	1.4				ug/L	15231	Standard
S	34	1995.1	3.9				ug/L	1845	Standard
Sr	88	96.7	26.5				ug/L	70	Standard
C	12	250.0	28.0				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	68.0	19.9				mg/L	12	Standard
Ho-1	165	86.7	18.5				mg/L	12	Standard
Er	166	180.0	33.3				mg/L	20	Standard
I	127	352295.0	10.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		131.653	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.768	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036207
Report Date/Time: Tuesday, June 28, 2016 14:08:19
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.512
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.748
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mo 98 Upper, S, EEE	Mo	98	
Cd 111 Lower	Cd	111	

Sample ID: L1606036207
Report Date/Time: Tuesday, June 28, 2016 14:08:19
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036208
Sample Date/Time: Tuesday, June 28, 2016 14:09:14
Number of Replicates: 3
Autosampler Position: 210
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	113878.0	0.7				ug/L	81760	Standard
	Be	9	15.0	33.3	**0.0011**	0.004	325.9	ug/L	17	Standard
	Al	27	36051659.4	1.3	**167.9026**	1.365	0.8	ug/L	1410	Standard
	Sc	45	40123.1	2.7				ug/L	39543	Standard
	Ti	47	157.7	1.6	**0.4795**	0.006	1.3	ug/L	39	Standard
	V	51	2374.4	7.5	**0.1409**	0.028	20.1	ug/L	1415	Standard
	Cr	52	8200.6	0.9	**0.1408**	0.017	11.9	ug/L	6946	Standard
	Cr	53	1531.7	7.2	**1.4329**	0.130	9.1	ug/L	328	Standard
	Mn	55	2760.9	2.4	**0.2845**	0.009	3.3	ug/L	823	Standard
	Co	59	1114.0	4.4	**0.1299**	0.007	5.8	ug/L	252	Standard
	Ni	60	3861.2	1.0	**1.9798**	0.033	1.7	ug/L	192	Standard
	Cu	65	1286.1	5.6	**0.5974**	0.039	6.5	ug/L	141	Standard
	Zn	66	2133.2	3.4	**1.1513**	0.069	6.0	ug/L	491	Standard
>	**Ge**	72	506573.0	0.6				ug/L	515087	Standard
	As	75	124.6	44.5	**0.1528**	0.050	32.6	ug/L	20	Standard
	Se	82	87.7	7.6	**0.6552**	0.055	8.5	ug/L	15	Standard
	Se-1	77	126.7	6.6	**0.9909**	0.116	11.7	ug/L	48	Standard
>	**Ga**	71	56.7	45.3				mg/L	90	Standard
	Rb	85	275.0	18.5				ug/L	23	Standard
	Y	89	450902.0	1.5				ug/L	448510	Standard
>	**Rh**	103	46.7	32.7				ug/L	15	Standard
	Mo	98	3181.2	4.8	**1.1379**	0.054	4.7	ug/L	11	Standard
	Ag	107	86.0	6.5	**-0.0036**	0.001	24.6	ug/L	134	Standard
	Cd	111	11.8	23.2	**0.0025**	0.001	50.9	mg/L	7	Standard
	Cd	114	54.8	27.0	**0.0054**	0.003	50.7	ug/L	16	Standard
>	**In**	115	621612.8	0.6				ug/L	611117	Standard
	Sn	118	878.4	13.7	**0.0442**	0.021	46.4	ug/L	518	Standard
	Sb	123	218.7	21.6	**0.0327**	0.010	30.0	ug/L	127	Standard
	Ba	135	24511.5	2.2	**11.6480**	0.297	2.5	ug/L	45	Standard
	Ce	140	126.7	17.8				ug/L	298	Standard
>	**Tb**	159	1020349.4	1.6				ug/L	977942	Standard
	Ho	165	38.3	41.9				ug/L	12	Standard
	Tl	203	70.7	5.4	**0.0034**	0.000	14.1	ug/L	31	Standard
	Tl	205	50.0	36.1	**0.0044**	0.002	54.4	ug/L	20	Standard
	Pb	206	339.3	6.0	**-0.0025**	0.003	133.5	ug/L	286	Standard
	Pb	207	287.3	3.0	**-0.0007**	0.002	279.9	ug/L	257	Standard
	Pb	208	1133.4	0.9	**-0.0020**	0.001	40.2	ug/L	956	Standard
	U	238	2248.5	2.1	**0.3023**	0.005	1.6	ug/L	17	Standard
>	**Bi**	209	572724.5	0.5				ug/L	557080	Standard

Sample ID: L1606036208
Report Date/Time: Tuesday, June 28, 2016 14:11:31
Page 1

Approved: June 29, 2016

Na	23	11.7	65.5	**74.1428**	49.752	67.1	mg/L	0	Standard
Mg	24	18449.6	1.2	**24.1624**	0.882	3.7	mg/L	32	Standard
K	39	318.3	12.8	**0.9203**	0.151	16.4	mg/L	30	Standard
Ca	43	671.7	8.0	**95.1609**	6.901	7.3	mg/L	48	Standard
Fe	54	172.3	21.8	**0.0165**	0.033	197.4	mg/L	159	Standard
Fe	57	548.3	9.0	**1.2014**	0.184	15.3	mg/L	195	Standard
> **Sc-1**	45	40123.1	2.7				mg/L	39543	Standard
Cl	35	45031.5	1.7				ug/L	27642	Standard
Kr	83	2.3	89.2				ug/L	2	Standard
Br	81	9216.1	3.8				ug/L	1647	Standard
P	31	22331.4	1.4				ug/L	15231	Standard
S	34	2088.5	6.2				ug/L	1845	Standard
Sr	88	246.7	15.5				ug/L	70	Standard
C	12	346.7	6.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	37.4	4.1				mg/L	12	Standard
Ho-1	165	38.3	41.9				mg/L	12	Standard
Er	166	53.3	60.3				mg/L	20	Standard
I	127	16496.5	38.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		139.283	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		98.347	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606036208
Report Date/Time: Tuesday, June 28, 2016 14:11:31
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.718
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.808
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036209
Sample Date/Time: Tuesday, June 28, 2016 14:12:25
Number of Replicates: 3
Autosampler Position: 211
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	107575.7	2.1				ug/L	81760	Standard
	Be	9	15.0	33.3	**0.0018**	0.004	223.3	ug/L	17	Standard
	Al	27	24368759.8	1.3	**120.1696**	2.653	2.2	ug/L	1410	Standard
	Sc	45	40144.8	3.7				ug/L	39543	Standard
	Ti	47	175.7	9.7	**0.5394**	0.064	11.9	ug/L	39	Standard
	V	51	1717.1	6.1	**0.0398**	0.013	33.2	ug/L	1415	Standard
	Cr	52	8804.6	0.4	**0.2048**	0.007	3.4	ug/L	6946	Standard
	Cr	53	1576.7	7.9	**1.4512**	0.154	10.6	ug/L	328	Standard
	Mn	55	1083613.1	0.9	**166.6059**	0.333	0.2	ug/L	823	Standard
	Co	59	2591.6	2.6	**0.3366**	0.007	2.1	ug/L	252	Standard
	Ni	60	4161.2	1.9	**2.0993**	0.030	1.4	ug/L	192	Standard
	Cu	65	1088.7	1.4	**0.4808**	0.011	2.3	ug/L	141	Standard
	Zn	66	3226.7	2.3	**2.1094**	0.068	3.2	ug/L	491	Standard
>	**Ge**	72	516304.1	0.8				ug/L	515087	Standard
	As	75	854.1	2.3	**0.7983**	0.019	2.4	ug/L	20	Standard
	Se	82	52.5	11.5	**0.3355**	0.056	16.6	ug/L	15	Standard
	Se-1	77	105.7	6.3	**0.6809**	0.097	14.3	ug/L	48	Standard
>	**Ga**	71	153.3	30.3				mg/L	90	Standard
	Rb	85	15286.0	2.4				ug/L	23	Standard
	Y	89	449805.6	1.9				ug/L	448510	Standard
>	**Rh**	103	36.7	43.8				ug/L	15	Standard
	Mo	98	1903.1	2.7	**0.6696**	0.018	2.6	ug/L	11	Standard
	Ag	107	88.0	6.3	**-0.0034**	0.001	30.6	ug/L	134	Standard
	Cd	111	24.4	20.6	**0.0083**	0.002	29.8	mg/L	7	Standard
	Cd	114	75.3	47.3	**0.0092**	0.007	71.3	ug/L	16	Standard
>	**In**	115	624865.1	1.8				ug/L	611117	Standard
	Sn	118	858.4	10.9	**0.0403**	0.018	44.8	ug/L	518	Standard
	Sb	123	598.6	8.7	**0.1114**	0.013	11.4	ug/L	127	Standard
	Ba	135	79235.0	1.1	**37.5150**	0.511	1.4	ug/L	45	Standard
	Ce	140	1043.4	5.0				ug/L	298	Standard
>	**Tb**	159	1006271.6	1.0				ug/L	977942	Standard
	Ho	165	31.7	55.5				ug/L	12	Standard
	Tl	203	94.7	10.4	**0.0060**	0.001	17.2	ug/L	31	Standard
	Tl	205	103.3	26.6	**0.0113**	0.003	31.1	ug/L	20	Standard
	Pb	206	1317.1	2.4	**0.1704**	0.008	4.5	ug/L	286	Standard
	Pb	207	1098.7	2.3	**0.1583**	0.006	3.8	ug/L	257	Standard
	Pb	208	4505.3	1.8	**0.1689**	0.006	3.3	ug/L	956	Standard
	U	238	1688.4	2.2	**0.2232**	0.003	1.5	ug/L	17	Standard
>	**Bi**	209	580367.6	0.9				ug/L	557080	Standard

Sample ID: L1606036209
Report Date/Time: Tuesday, June 28, 2016 14:14:42
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**32.4402**	32.912	101.5	mg/L	0	Standard
Mg	24	8242.2	2.3	**10.7549**	0.236	2.2	mg/L	32	Standard
K	39	375.0	23.7	**1.1102**	0.318	28.6	mg/L	30	Standard
Ca	43	428.3	18.6	**58.5393**	9.751	16.7	mg/L	48	Standard
Fe	54	1288.6	9.8	**1.0145**	0.141	13.9	mg/L	159	Standard
Fe	57	713.4	8.9	**1.7461**	0.160	9.2	mg/L	195	Standard
> **Sc-1**	45	40144.8	3.7				mg/L	39543	Standard
Cl	35	37853.9	1.3				ug/L	27642	Standard
Kr	83	2.3	173.2				ug/L	2	Standard
Br	81	8949.3	3.3				ug/L	1647	Standard
P	31	22830.5	2.4				ug/L	15231	Standard
S	34	2348.5	6.0				ug/L	1845	Standard
Sr	88	131.7	18.7				ug/L	70	Standard
C	12	320.0	23.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	57.1	19.2				mg/L	12	Standard
Ho-1	165	31.7	55.5				mg/L	12	Standard
Er	166	60.0	44.1				mg/L	20	Standard
I	127	12798.7	0.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		131.575	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.236	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036209
Report Date/Time: Tuesday, June 28, 2016 14:14:42
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	102.250
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.180
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606036209
Report Date/Time: Tuesday, June 28, 2016 14:14:42
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036210
Sample Date/Time: Tuesday, June 28, 2016 14:15:37
Number of Replicates: 3
Autosampler Position: 212
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	112279.5	1.4				ug/L	81760	Standard
	Be	9	23.3	32.7	**0.0074**	0.006	75.4	ug/L	17	Standard
	Al	27	11519588.8	0.5	**54.4098**	0.542	1.0	ug/L	1410	Standard
	Sc	45	38345.1	1.2				ug/L	39543	Standard
	Ti	47	94.7	7.9	**0.2480**	0.026	10.6	ug/L	39	Standard
	V	51	2217.7	3.9	**0.1400**	0.014	10.3	ug/L	1415	Standard
	Cr	52	7580.6	0.1	**0.1258**	0.020	16.0	ug/L	6946	Standard
	Cr	53	2436.9	5.1	**2.6754**	0.201	7.5	ug/L	328	Standard
	Mn	55	75388.8	2.2	**12.4826**	0.333	2.7	ug/L	823	Standard
	Co	59	4144.2	3.9	**0.6094**	0.028	4.6	ug/L	252	Standard
	Ni	60	4832.4	1.6	**2.6820**	0.049	1.8	ug/L	192	Standard
	Cu	65	4625.4	2.6	**2.5532**	0.075	2.9	ug/L	141	Standard
	Zn	66	4203.9	2.1	**3.3383**	0.098	2.9	ug/L	491	Standard
>	**Ge**	72	474293.4	1.6				ug/L	515087	Standard
	As	75	1073.0	2.0	**1.0772**	0.038	3.5	ug/L	20	Standard
	Se	82	351.4	3.1	**3.1960**	0.130	4.1	ug/L	15	Standard
	Se-1	77	210.7	7.1	**2.3175**	0.198	8.5	ug/L	48	Standard
>	**Ga**	71	73.3	33.6				mg/L	90	Standard
	Rb	85	56526.9	1.1				ug/L	23	Standard
	Y	89	421082.0	1.6				ug/L	448510	Standard
>	**Rh**	103	96.7	14.9				ug/L	15	Standard
	Mo	98	257904.5	1.0	**100.5156**	1.205	1.2	ug/L	11	Standard
	Ag	107	79.7	12.8	**-0.0037**	0.002	45.1	ug/L	134	Standard
	Cd	111	-189.4	16.3	**-0.0962**	0.015	15.7	mg/L	7	Standard
	Cd	114	358.5	16.9	**0.0677**	0.012	17.5	ug/L	16	Standard
>	**In**	115	580013.8	0.6				ug/L	611117	Standard
	Sn	118	865.0	5.3	**0.0522**	0.009	16.9	ug/L	518	Standard
	Sb	123	234.3	5.5	**0.0395**	0.003	7.3	ug/L	127	Standard
	Ba	135	12298.0	0.4	**6.2513**	0.035	0.6	ug/L	45	Standard
	Ce	140	386.7	8.8				ug/L	298	Standard
>	**Tb**	159	979000.4	0.2				ug/L	977942	Standard
	Ho	165	86.7	26.6				ug/L	12	Standard
	Tl	203	556.3	8.3	**0.0635**	0.006	10.1	ug/L	31	Standard
	Tl	205	491.7	8.6	**0.0672**	0.007	10.1	ug/L	20	Standard
	Pb	206	339.7	4.4	**0.0013**	0.002	176.0	ug/L	286	Standard
	Pb	207	303.3	2.7	**0.0062**	0.001	17.9	ug/L	257	Standard
	Pb	208	1090.3	1.9	**-0.0007**	0.002	299.6	ug/L	956	Standard
	U	238	117.3	12.8	**0.0140**	0.002	17.0	ug/L	17	Standard
>	**Bi**	209	539461.3	1.8				ug/L	557080	Standard

Sample ID: L1606036210
Report Date/Time: Tuesday, June 28, 2016 14:17:54
Page 1

Approved: June 29, 2016

Na	23	8.3	34.6	**55.0323**	18.599	33.8	mg/L	0	Standard
Mg	24	255159.1	1.2	**350.2417**	6.027	1.7	mg/L	32	Standard
K	39	5237.6	3.5	**18.1047**	0.448	2.5	mg/L	30	Standard
Ca	43	815.0	11.2	**122.4307**	15.864	13.0	mg/L	48	Standard
Fe	54	150.7	21.1	**0.0038**	0.031	813.1	mg/L	159	Standard
Fe	57	818.4	11.3	**2.2227**	0.299	13.5	mg/L	195	Standard
> **Sc-1**	45	38345.1	1.2				mg/L	39543	Standard
Cl	35	84049.9	1.6				ug/L	27642	Standard
Kr	83	2.0	100.0				ug/L	2	Standard
Br	81	57144.3	3.2				ug/L	1647	Standard
P	31	18494.6	4.2				ug/L	15231	Standard
S	34	1918.5	8.3				ug/L	1845	Standard
Sr	88	445.0	8.9				ug/L	70	Standard
C	12	253.3	16.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	64.7	80.3				mg/L	12	Standard
Ho-1	165	86.7	26.6				mg/L	12	Standard
Er	166	110.0	9.1				mg/L	20	Standard
I	127	999438.0	9.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		137.328	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.080	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036210
Report Date/Time: Tuesday, June 28, 2016 14:17:54
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.911
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.837
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mo 98 Upper, S, EEE	Mo	98	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 14:18:50
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	108145.2	5.1				ug/L	81760	Standard
	Be	9	64242.8	4.5	**48.9113**	0.992	2.0	ug/L	17	Standard
	Al	27	8304059.2	3.3	**40.7429**	1.039	2.5	ug/L	1410	Standard
	Sc	45	37765.4	4.0				ug/L	39543	Standard
	Ti	47	23471.8	2.0	**97.1021**	1.269	1.3	ug/L	39	Standard
	V	51	349107.5	0.6	**51.5604**	0.772	1.5	ug/L	1415	Standard
	Cr	52	335198.9	1.2	**48.5745**	0.974	2.0	ug/L	6946	Standard
	Cr	53	41869.4	1.3	**48.8267**	1.454	3.0	ug/L	328	Standard
	Mn	55	314993.6	1.3	**49.9295**	0.954	1.9	ug/L	823	Standard
	Co	59	375345.6	0.1	**55.0339**	1.228	2.2	ug/L	252	Standard
	Ni	60	88766.6	0.5	**48.4746**	0.838	1.7	ug/L	192	Standard
	Cu	65	90675.9	1.2	**49.0916**	0.682	1.4	ug/L	141	Standard
	Zn	66	53067.8	0.6	**49.0759**	0.955	1.9	ug/L	491	Standard
>	Ge	72	499870.2	2.1				ug/L	515087	Standard
	As	75	55072.1	0.8	**50.5344**	1.449	2.9	ug/L	20	Standard
	Se	82	6022.1	1.2	**53.7900**	0.803	1.5	ug/L	15	Standard
	Se-1	77	3835.5	2.7	**51.7480**	1.296	2.5	ug/L	48	Standard
>	Ga	71	68.3	34.6				mg/L	90	Standard
	Rb	85	878.4	4.0				ug/L	23	Standard
	Y	89	444195.9	1.0				ug/L	448510	Standard
>	Rh	103	36.7	7.9				ug/L	15	Standard
	Mo	98	271765.3	0.4	**101.0586**	1.139	1.1	ug/L	11	Standard
	Ag	107	335937.1	0.8	**50.2935**	0.228	0.5	ug/L	134	Standard
	Cd	111	104881.4	1.9	**49.3201**	0.733	1.5	mg/L	7	Standard
	Cd	114	258223.7	2.1	**49.9610**	1.276	2.6	ug/L	16	Standard
>	In	115	607920.9	0.8				ug/L	611117	Standard
	Sn	118	297159.7	1.8	**50.4160**	0.583	1.2	ug/L	518	Standard
	Sb	123	233566.6	0.5	**49.7525**	0.427	0.9	ug/L	127	Standard
	Ba	135	100029.2	1.5	**48.6806**	0.509	1.0	ug/L	45	Standard
	Ce	140	56.7	27.0				ug/L	298	Standard
>	Tb	159	990459.3	1.9				ug/L	977942	Standard
	Ho	165	13.3	21.7				ug/L	12	Standard
	Tl	203	438180.8	0.9	**49.9571**	0.368	0.7	ug/L	31	Standard
	Tl	205	400912.4	1.6	**52.6337**	0.214	0.4	ug/L	20	Standard
	Pb	206	285968.6	0.8	**50.7733**	0.808	1.6	ug/L	286	Standard
	Pb	207	258378.5	0.4	**50.8616**	0.645	1.3	ug/L	257	Standard
	Pb	208	1010283.1	0.5	**51.4021**	0.498	1.0	ug/L	956	Standard
	U	238	389330.8	1.6	**52.1963**	0.349	0.7	ug/L	17	Standard
>	Bi	209	579902.5	1.4				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:21:06
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4220.6	3.3	**5.8326**	0.328	5.6	mg/L	32	Standard
K	39	1343.4	3.1	**4.6154**	0.216	4.7	mg/L	30	Standard
Ca	43	66.7	37.7	**5.1699**	3.623	70.1	mg/L	48	Standard
Fe	54	5699.2	5.0	**5.2644**	0.205	3.9	mg/L	159	Standard
Fe	57	1620.1	3.2	**5.1052**	0.186	3.6	mg/L	195	Standard
> **Sc-1**	45	37765.4	4.0				mg/L	39543	Standard
Cl	35	31900.9	3.3				ug/L	27642	Standard
Kr	83	2.3	89.2				ug/L	2	Standard
Br	81	2500.2	16.5				ug/L	1647	Standard
P	31	18272.7	3.1				ug/L	15231	Standard
S	34	2018.5	1.5				ug/L	1845	Standard
Sr	88	43.3	24.0				ug/L	70	Standard
C	12	196.7	21.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	12.4	50.3				mg/L	12	Standard
Ho-1	165	13.3	21.7				mg/L	12	Standard
Er	166	20.0	86.6				mg/L	20	Standard
I	127	32500.9	63.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	97.823		
Al	27	81.486		
Sc	45			
Ti	47	97.102		
V	51	103.121		
Cr	52	97.149		
Cr	53			
Mn	55	99.859		
Co	59	110.068		
Ni	60	96.949		
Cu	65	98.183		
Zn	66	98.152		
> Ge	72		97.046	
As	75	101.069		
Se	82	107.580		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:21:06
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.059		
Ag	107	100.587		
Cd	111	98.640		
Cd	114			
In	115		99.477	
Sn	118	100.832		
Sb	123	99.505		
Ba	135	97.361		
Ce	140			
Tb	159			
Ho	165			
Tl	203	99.914		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.804		
U	238	104.393		
Bi	209		104.097	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:21:06
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 14:22:01
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101364.8	2.9				ug/L	81760	Standard
	Be	9	38.3	49.4	**0.0216**	0.016	74.7	ug/L	17	Standard
	Al	27	15489.3	94.8	**0.0675**	0.079	117.4	ug/L	1410	Standard
	Sc	45	38064.4	0.9				ug/L	39543	Standard
	Ti	47	37.3	29.4	**-0.0111**	0.045	403.8	ug/L	39	Standard
	V	51	1161.3	1.6	**-0.0351**	0.002	6.4	ug/L	1415	Standard
	Cr	52	5982.8	2.5	**-0.1747**	0.020	11.4	ug/L	6946	Standard
	Cr	53	318.3	7.1	**0.0308**	0.025	81.8	ug/L	328	Standard
	Mn	55	894.7	9.5	**-0.0068**	0.013	197.2	ug/L	823	Standard
	Co	59	241.0	14.9	**0.0039**	0.005	138.8	ug/L	252	Standard
	Ni	60	266.0	10.4	**0.0400**	0.015	37.4	ug/L	192	Standard
	Cu	65	200.0	8.9	**0.0172**	0.009	54.5	ug/L	141	Standard
	Zn	66	430.0	7.9	**-0.4249**	0.030	7.1	ug/L	491	Standard
>	**Ge**	72	501700.0	0.4				ug/L	515087	Standard
	As	75	1.5	1379.3	**0.0417**	0.019	44.5	ug/L	20	Standard
	Se	82	27.3	19.7	**0.1242**	0.049	39.3	ug/L	15	Standard
	Se-1	77	54.3	18.4	**0.0219**	0.139	635.7	ug/L	48	Standard
>	**Ga**	71	63.3	9.1				mg/L	90	Standard
	Rb	85	26.7	28.6				ug/L	23	Standard
	Y	89	434979.7	1.5				ug/L	448510	Standard
>	**Rh**	103	11.7	89.2				ug/L	15	Standard
	Mo	98	412.1	20.1	**0.1325**	0.026	19.5	ug/L	11	Standard
	Ag	107	133.7	24.7	**0.0036**	0.005	131.9	ug/L	134	Standard
	Cd	111	17.6	36.9	**0.0053**	0.003	55.6	mg/L	7	Standard
	Cd	114	37.0	22.5	**0.0022**	0.002	81.0	ug/L	16	Standard
>	**In**	115	612643.0	3.0				ug/L	611117	Standard
	Sn	118	796.7	19.5	**0.0328**	0.028	85.4	ug/L	518	Standard
	Sb	123	465.9	23.7	**0.0860**	0.025	29.2	ug/L	127	Standard
	Ba	135	55.7	25.4	**0.0018**	0.007	380.6	ug/L	45	Standard
	Ce	140	23.3	53.9				ug/L	298	Standard
>	**Tb**	159	991657.2	2.1				ug/L	977942	Standard
	Ho	165	6.7	114.6				ug/L	12	Standard
	Tl	203	81.3	30.4	**0.0043**	0.003	63.5	ug/L	31	Standard
	Tl	205	65.0	13.3	**0.0061**	0.001	17.5	ug/L	20	Standard
	Pb	206	313.0	2.4	**-0.0091**	0.002	17.1	ug/L	286	Standard
	Pb	207	268.3	2.3	**-0.0063**	0.001	16.7	ug/L	257	Standard
	Pb	208	1057.3	6.3	**-0.0078**	0.003	38.5	ug/L	956	Standard
	U	238	61.3	40.7	**0.0051**	0.003	63.9	ug/L	17	Standard
>	**Bi**	209	592397.8	0.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 14:24:18
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	95.0	63.8	**0.0743**	0.084	112.6	mg/L	32	Standard
K	39	16.7	17.3	**-0.0829**	0.011	12.9	mg/L	30	Standard
Ca	43	31.7	36.5	**-0.3389**	1.874	552.9	mg/L	48	Standard
Fe	54	110.8	32.6	**-0.0331**	0.033	100.8	mg/L	159	Standard
Fe	57	168.3	28.5	**-0.0341**	0.169	494.5	mg/L	195	Standard
> **Sc-1**	45	38064.4	0.9				mg/L	39543	Standard
Cl	35	31593.6	0.5				ug/L	27642	Standard
Kr	83	0.3	173.2				ug/L	2	Standard
Br	81	1856.8	9.2				ug/L	1647	Standard
P	31	17672.0	1.1				ug/L	15231	Standard
S	34	2055.1	6.3				ug/L	1845	Standard
Sr	88	56.7	18.4				ug/L	70	Standard
C	12	170.0	20.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	6.3	173.2				mg/L	12	Standard
Ho-1	165	6.7	114.6				mg/L	12	Standard
Er	166	6.7	173.2				mg/L	20	Standard
I	127	7793.7	20.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.401	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 14:24:18
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.250
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.340
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036211
Sample Date/Time: Tuesday, June 28, 2016 14:25:14
Number of Replicates: 3
Autosampler Position: 213
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100807.6	1.2				ug/L	81760	Standard
	Be	9	76.7	13.6	**0.0528**	0.008	15.0	ug/L	17	Standard
	Al	27	23829161.3	1.5	**125.3655**	0.841	0.7	ug/L	1410	Standard
	Sc	45	39760.5	3.7				ug/L	39543	Standard
	Ti	47	180.7	0.8	**0.5860**	0.014	2.4	ug/L	39	Standard
	V	51	2583.8	5.2	**0.1775**	0.016	9.1	ug/L	1415	Standard
	Cr	52	8684.5	0.7	**0.2325**	0.005	2.3	ug/L	6946	Standard
	Cr	53	1498.4	3.7	**1.4227**	0.048	3.4	ug/L	328	Standard
	Mn	55	44479.2	1.1	**6.9441**	0.116	1.7	ug/L	823	Standard
	Co	59	1598.1	1.9	**0.2038**	0.006	2.9	ug/L	252	Standard
	Ni	60	3161.3	2.5	**1.6303**	0.047	2.9	ug/L	192	Standard
	Cu	65	697.7	3.7	**0.2888**	0.018	6.1	ug/L	141	Standard
	Zn	66	2205.8	4.6	**1.2527**	0.096	7.7	ug/L	491	Standard
>	**Ge**	72	498311.0	1.0				ug/L	515087	Standard
	As	75	175.3	10.2	**0.2014**	0.015	7.3	ug/L	20	Standard
	Se	82	90.2	6.0	**0.6901**	0.041	6.0	ug/L	15	Standard
	Se-1	77	104.3	15.6	**0.7141**	0.236	33.0	ug/L	48	Standard
>	**Ga**	71	91.7	19.2				mg/L	90	Standard
	Rb	85	3617.1	1.0				ug/L	23	Standard
	Y	89	438807.3	1.1				ug/L	448510	Standard
>	**Rh**	103	23.3	24.7				ug/L	15	Standard
	Mo	98	5340.3	1.5	**1.9620**	0.035	1.8	ug/L	11	Standard
	Ag	107	313.7	8.5	**0.0307**	0.004	14.3	ug/L	134	Standard
	Cd	111	102.5	10.0	**0.0451**	0.004	9.9	mg/L	7	Standard
	Cd	114	269.5	9.6	**0.0471**	0.005	11.0	ug/L	16	Standard
>	**In**	115	609490.3	1.0				ug/L	611117	Standard
	Sn	118	888.4	6.3	**0.0488**	0.011	22.0	ug/L	518	Standard
	Sb	123	739.4	5.4	**0.1443**	0.009	6.4	ug/L	127	Standard
	Ba	135	23651.8	1.3	**11.4620**	0.117	1.0	ug/L	45	Standard
	Ce	140	2515.2	8.7				ug/L	298	Standard
>	**Tb**	159	994391.3	1.1				ug/L	977942	Standard
	Ho	165	61.7	16.9				ug/L	12	Standard
	Tl	203	495.7	2.3	**0.0524**	0.001	2.2	ug/L	31	Standard
	Tl	205	451.7	6.1	**0.0577**	0.003	5.9	ug/L	20	Standard
	Pb	206	782.0	4.2	**0.0770**	0.005	7.0	ug/L	286	Standard
	Pb	207	638.7	5.2	**0.0693**	0.006	8.1	ug/L	257	Standard
	Pb	208	2612.4	2.6	**0.0742**	0.003	3.5	ug/L	956	Standard
	U	238	3078.0	1.8	**0.4146**	0.006	1.5	ug/L	17	Standard
>	**Bi**	209	573149.3	0.8				ug/L	557080	Standard

Sample ID: L1606036211
Report Date/Time: Tuesday, June 28, 2016 14:27:31
Page 1

Approved: June 29, 2016

Na	23	10.0	50.0	**63.1287**	29.742	47.1	mg/L	0	Standard
Mg	24	8182.2	4.0	**10.7839**	0.556	5.2	mg/L	32	Standard
K	39	235.0	11.8	**0.6477**	0.082	12.7	mg/L	30	Standard
Ca	43	488.3	8.8	**68.4110**	5.631	8.2	mg/L	48	Standard
Fe	54	190.4	22.3	**0.0341**	0.036	104.7	mg/L	159	Standard
Fe	57	451.7	3.2	**0.8921**	0.019	2.1	mg/L	195	Standard
Sc-1	45	39760.5	3.7				mg/L	39543	Standard
Cl	35	36572.1	2.1				ug/L	27642	Standard
Kr	83	2.0	0.0				ug/L	2	Standard
Br	81	15501.3	5.9				ug/L	1647	Standard
P	31	22299.7	2.1				ug/L	15231	Standard
S	34	2183.5	4.1				ug/L	1845	Standard
Sr	88	141.7	11.3				ug/L	70	Standard
C	12	286.7	23.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	128.1	22.2				mg/L	12	Standard
Ho-1	165	61.7	16.9				mg/L	12	Standard
Er	166	110.0	63.6				mg/L	20	Standard
I	127	9539.7	2.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		123.297	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.743	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036211
Report Date/Time: Tuesday, June 28, 2016 14:27:31
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	99.734
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.884
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606036211
Report Date/Time: Tuesday, June 28, 2016 14:27:31
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036212
Sample Date/Time: Tuesday, June 28, 2016 14:28:25
Number of Replicates: 3
Autosampler Position: 214
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101556.0	2.3				ug/L	81760	Standard
	Be	9	16.7	34.6	**0.0037**	0.005	121.7	ug/L	17	Standard
	Al	27	17765744.4	2.3	**92.8094**	3.109	3.3	ug/L	1410	Standard
	Sc	45	39087.0	2.2				ug/L	39543	Standard
	Ti	47	247.0	6.3	**0.8686**	0.057	6.6	ug/L	39	Standard
	V	51	2382.1	6.3	**0.1501**	0.021	13.9	ug/L	1415	Standard
	Cr	52	8300.6	1.1	**0.1842**	0.008	4.2	ug/L	6946	Standard
	Cr	53	1501.7	3.0	**1.4394**	0.046	3.2	ug/L	328	Standard
	Mn	55	1258376.0	0.7	**201.9190**	0.817	0.4	ug/L	823	Standard
	Co	59	3356.7	1.5	**0.4660**	0.012	2.5	ug/L	252	Standard
	Ni	60	3550.4	0.9	**1.8576**	0.023	1.3	ug/L	192	Standard
	Cu	65	1089.7	4.5	**0.5061**	0.022	4.4	ug/L	141	Standard
	Zn	66	2532.2	4.9	**1.5769**	0.099	6.3	ug/L	491	Standard
>	**Ge**	72	494799.1	0.8				ug/L	515087	Standard
	As	75	371.2	4.5	**0.3840**	0.014	3.7	ug/L	20	Standard
	Se	82	96.6	14.9	**0.7542**	0.127	16.9	ug/L	15	Standard
	Se-1	77	117.3	12.8	**0.9020**	0.200	22.2	ug/L	48	Standard
>	**Ga**	71	161.7	20.6				mg/L	90	Standard
	Rb	85	5996.2	5.2				ug/L	23	Standard
	Y	89	446186.9	1.4				ug/L	448510	Standard
>	**Rh**	103	35.0	37.8				ug/L	15	Standard
	Mo	98	1982.4	1.7	**0.7175**	0.021	3.0	ug/L	11	Standard
	Ag	107	87.7	6.7	**-0.0031**	0.001	30.7	ug/L	134	Standard
	Cd	111	81.3	17.6	**0.0353**	0.007	20.5	mg/L	7	Standard
	Cd	114	230.7	12.0	**0.0397**	0.006	14.9	ug/L	16	Standard
>	**In**	115	608659.5	1.3				ug/L	611117	Standard
	Sn	118	906.7	3.5	**0.0521**	0.007	12.9	ug/L	518	Standard
	Sb	123	394.5	7.6	**0.0711**	0.006	8.6	ug/L	127	Standard
	Ba	135	51224.8	1.3	**24.8890**	0.366	1.5	ug/L	45	Standard
	Ce	140	3013.6	4.4				ug/L	298	Standard
>	**Tb**	159	974496.5	1.8				ug/L	977942	Standard
	Ho	165	80.0	16.5				ug/L	12	Standard
	Tl	203	128.3	8.3	**0.0104**	0.001	10.4	ug/L	31	Standard
	Tl	205	91.7	22.0	**0.0102**	0.003	28.0	ug/L	20	Standard
	Pb	206	755.4	5.1	**0.0757**	0.008	10.8	ug/L	286	Standard
	Pb	207	644.0	3.0	**0.0736**	0.005	7.3	ug/L	257	Standard
	Pb	208	2468.4	2.8	**0.0700**	0.005	7.5	ug/L	956	Standard
	U	238	1502.1	3.5	**0.2059**	0.006	2.8	ug/L	17	Standard
>	**Bi**	209	559159.1	1.2				ug/L	557080	Standard

Sample ID: L1606036212
Report Date/Time: Tuesday, June 28, 2016 14:30:42
Page 1

Approved: June 29, 2016

Na	23	5.0	0.0	**32.4897**	0.713	2.2	mg/L	0	Standard
Mg	24	34367.3	2.9	**46.2225**	0.714	1.5	mg/L	32	Standard
K	39	518.3	10.6	**1.6321**	0.206	12.7	mg/L	30	Standard
Ca	43	333.3	3.5	**45.8704**	1.237	2.7	mg/L	48	Standard
Fe	54	313.8	12.6	**0.1508**	0.042	28.0	mg/L	159	Standard
Fe	57	493.3	3.3	**1.0616**	0.089	8.4	mg/L	195	Standard
> **Sc-1**	45	39087.0	2.2				mg/L	39543	Standard
Cl	35	40128.4	3.1				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	13105.7	7.2				ug/L	1647	Standard
P	31	21068.0	1.2				ug/L	15231	Standard
S	34	2031.8	3.8				ug/L	1845	Standard
Sr	88	195.0	14.3				ug/L	70	Standard
C	12	220.0	19.8				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	113.5	27.0				mg/L	12	Standard
Ho-1	165	80.0	16.5				mg/L	12	Standard
Er	166	66.7	48.2				mg/L	20	Standard
I	127	14735.5	4.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		124.212	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.061	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036212
Report Date/Time: Tuesday, June 28, 2016 14:30:42
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.598
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.373
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 14:31:38
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96993.2	4.5				ug/L	81760	Standard
	Be	9	59018.4	6.5	**50.0965**	2.769	5.5	ug/L	17	Standard
	Al	27	7977258.4	4.2	**43.6214**	1.029	2.4	ug/L	1410	Standard
	Sc	45	38238.2	1.6				ug/L	39543	Standard
	Ti	47	23612.4	4.0	**98.7594**	3.955	4.0	ug/L	39	Standard
	V	51	340072.0	1.7	**50.7725**	1.247	2.5	ug/L	1415	Standard
	Cr	52	338200.8	3.5	**49.5590**	1.660	3.3	ug/L	6946	Standard
	Cr	53	42016.5	3.2	**49.5286**	1.703	3.4	ug/L	328	Standard
	Mn	55	307360.3	1.5	**49.2454**	0.735	1.5	ug/L	823	Standard
	Co	59	358545.3	2.7	**53.1356**	1.497	2.8	ug/L	252	Standard
	Ni	60	89705.8	2.1	**49.5206**	1.132	2.3	ug/L	192	Standard
	Cu	65	91128.3	2.1	**49.8745**	1.003	2.0	ug/L	141	Standard
	Zn	66	53559.8	1.4	**50.0843**	0.931	1.9	ug/L	491	Standard
>	**Ge**	72	494423.7	1.0				ug/L	515087	Standard
	As	75	53994.0	0.9	**50.0733**	0.496	1.0	ug/L	20	Standard
	Se	82	5715.9	0.2	**51.6065**	0.476	0.9	ug/L	15	Standard
	Se-1	77	3736.8	1.4	**50.9602**	0.952	1.9	ug/L	48	Standard
>	**Ga**	71	70.0	28.6				mg/L	90	Standard
	Rb	85	770.0	3.6				ug/L	23	Standard
	Y	89	445528.9	0.6				ug/L	448510	Standard
>	**Rh**	103	23.3	44.6				ug/L	15	Standard
	Mo	98	264652.3	1.1	**96.1672**	0.791	0.8	ug/L	11	Standard
	Ag	107	331013.3	2.8	**48.4244**	1.155	2.4	ug/L	134	Standard
	Cd	111	104542.0	2.3	**48.0434**	1.120	2.3	mg/L	7	Standard
	Cd	114	255433.5	4.1	**48.2912**	1.921	4.0	ug/L	16	Standard
>	**In**	115	622075.2	0.5				ug/L	611117	Standard
	Sn	118	295472.2	1.9	**48.9864**	0.743	1.5	ug/L	518	Standard
	Sb	123	230757.9	1.7	**48.0319**	0.618	1.3	ug/L	127	Standard
	Ba	135	99769.8	2.2	**47.4498**	0.988	2.1	ug/L	45	Standard
	Ce	140	73.3	7.9				ug/L	298	Standard
>	**Tb**	159	1000112.3	1.5				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	427237.6	2.4	**49.0359**	1.247	2.5	ug/L	31	Standard
	Tl	205	388094.9	2.7	**51.3011**	1.717	3.3	ug/L	20	Standard
	Pb	206	274003.7	2.0	**48.9721**	1.261	2.6	ug/L	286	Standard
	Pb	207	249924.4	1.5	**49.5263**	1.133	2.3	ug/L	257	Standard
	Pb	208	988217.1	0.7	**50.6156**	0.799	1.6	ug/L	956	Standard
	U	238	380024.3	2.2	**51.3005**	1.621	3.2	ug/L	17	Standard
>	**Bi**	209	576058.0	1.3				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:33:55
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**33.2699**	32.995	99.2	mg/L	0	Standard
Mg	24	4203.9	5.4	**5.7315**	0.340	5.9	mg/L	32	Standard
K	39	1220.0	6.7	**4.1203**	0.250	6.1	mg/L	30	Standard
Ca	43	103.3	14.8	**10.9126**	2.675	24.5	mg/L	48	Standard
Fe	54	5915.4	3.1	**5.3996**	0.173	3.2	mg/L	159	Standard
Fe	57	1661.8	2.4	**5.1766**	0.085	1.6	mg/L	195	Standard
> **Sc-1**	45	38238.2	1.6				mg/L	39543	Standard
Cl	35	33728.9	2.6				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	1893.5	8.5				ug/L	1647	Standard
P	31	18936.8	1.0				ug/L	15231	Standard
S	34	2451.9	4.7				ug/L	1845	Standard
Sr	88	73.3	14.2				ug/L	70	Standard
C	12	160.0	43.3				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.2	6.0				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	16.7	69.3				mg/L	20	Standard
I	127	3007.0	16.4				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	100.193		
	Al	27	87.243		
	Sc	45			
	Ti	47	98.759		
	V	51	101.545		
	Cr	52	99.118		
	Cr	53			
	Mn	55	98.491		
	Co	59	106.271		
	Ni	60	99.041		
	Cu	65	99.749		
	Zn	66	100.169		
>	Ge	72		95.988	
	As	75	100.147		
	Se	82	103.213		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:33:55
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	96.167		
Ag	107	96.849		
Cd	111	96.087		
Cd	114			
In	115			101.793
Sn	118	97.973		
Sb	123	96.064		
Ba	135	94.900		
Ce	140			
Tb	159			
Ho	165			
Tl	203	98.072		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.231		
U	238	102.601		
Bi	209			103.407
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:33:55
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 14:34:50
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92870.5	2.0				ug/L	81760	Standard
	Be	9	15.0	66.7	**0.0036**	0.009	245.7	ug/L	17	Standard
	Al	27	2743.6	21.7	**0.0013**	0.004	284.3	ug/L	1410	Standard
	Sc	45	38460.4	1.3				ug/L	39543	Standard
	Ti	47	30.7	8.2	**-0.0403**	0.010	23.8	ug/L	39	Standard
	V	51	1134.9	2.6	**-0.0414**	0.005	11.0	ug/L	1415	Standard
	Cr	52	5878.1	2.2	**-0.2023**	0.026	12.8	ug/L	6946	Standard
	Cr	53	310.0	1.6	**0.0160**	0.005	29.1	ug/L	328	Standard
	Mn	55	709.4	9.1	**-0.0377**	0.010	26.9	ug/L	823	Standard
	Co	59	233.7	1.1	**0.0023**	0.001	24.5	ug/L	252	Standard
	Ni	60	211.0	9.6	**0.0084**	0.011	126.5	ug/L	192	Standard
	Cu	65	155.3	8.6	**-0.0081**	0.008	96.4	ug/L	141	Standard
	Zn	66	394.0	4.1	**-0.4638**	0.014	3.1	ug/L	491	Standard
>	**Ge**	72	508932.6	0.9				ug/L	515087	Standard
	As	75	-34.1	13.4	**0.0096**	0.004	45.1	ug/L	20	Standard
	Se	82	19.4	20.9	**0.0515**	0.036	69.8	ug/L	15	Standard
	Se-1	77	46.7	12.2	**-0.0918**	0.077	84.0	ug/L	48	Standard
>	**Ga**	71	66.7	8.7				mg/L	90	Standard
	Rb	85	15.0	33.3				ug/L	23	Standard
	Y	89	445570.9	1.5				ug/L	448510	Standard
>	**Rh**	103	15.0	88.2				ug/L	15	Standard
	Mo	98	287.3	16.9	**0.0864**	0.017	20.0	ug/L	11	Standard
	Ag	107	124.0	22.6	**0.0021**	0.004	200.7	ug/L	134	Standard
	Cd	111	14.8	9.5	**0.0040**	0.001	16.6	mg/L	7	Standard
	Cd	114	48.0	37.7	**0.0043**	0.004	82.8	ug/L	16	Standard
>	**In**	115	615541.5	0.9				ug/L	611117	Standard
	Sn	118	883.4	22.9	**0.0465**	0.035	74.9	ug/L	518	Standard
	Sb	123	479.5	25.7	**0.0882**	0.027	30.1	ug/L	127	Standard
	Ba	135	35.7	10.6	**-0.0080**	0.002	22.2	ug/L	45	Standard
	Ce	140	21.7	48.0				ug/L	298	Standard
>	**Tb**	159	980265.4	0.4				ug/L	977942	Standard
	Ho	165	16.7	45.8				ug/L	12	Standard
	Tl	203	63.0	23.1	**0.0024**	0.002	67.2	ug/L	31	Standard
	Tl	205	63.3	31.9	**0.0060**	0.003	42.7	ug/L	20	Standard
	Pb	206	313.3	5.4	**-0.0080**	0.003	36.6	ug/L	286	Standard
	Pb	207	273.0	6.7	**-0.0043**	0.004	87.0	ug/L	257	Standard
	Pb	208	1087.7	7.3	**-0.0051**	0.004	85.3	ug/L	956	Standard
	U	238	49.0	30.5	**0.0036**	0.002	56.1	ug/L	17	Standard
>	**Bi**	209	580619.2	1.2				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 14:37:07
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	35.0	28.6	**-0.0091**	0.014	156.5	mg/L	32	Standard
K	39	36.7	28.4	**-0.0142**	0.035	248.1	mg/L	30	Standard
Ca	43	26.7	57.3	**-1.1996**	2.339	195.0	mg/L	48	Standard
Fe	54	140.8	15.9	**-0.0061**	0.021	341.9	mg/L	159	Standard
Fe	57	151.7	6.9	**-0.0978**	0.042	43.3	mg/L	195	Standard
> **Sc-1**	45	38460.4	1.3				mg/L	39543	Standard
Cl	35	32500.2	3.2				ug/L	27642	Standard
Kr	83	3.0	0.0				ug/L	2	Standard
Br	81	1736.8	3.3				ug/L	1647	Standard
P	31	17698.7	3.2				ug/L	15231	Standard
S	34	2320.2	10.5				ug/L	1845	Standard
Sr	88	73.3	21.9				ug/L	70	Standard
C	12	133.3	15.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	9.0	113.3				mg/L	12	Standard
Ho-1	165	16.7	45.8				mg/L	12	Standard
Er	166	20.0	50.0				mg/L	20	Standard
I	127	2206.8	20.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.805	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 14:37:07
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.724
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.225
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 14:37:07
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: PBS M7 WG573333-03
Sample Date/Time: Tuesday, June 28, 2016 14:38:03
Number of Replicates: 3
Autosampler Position: 215
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95050.7	3.1				ug/L	81760	Standard
	Be	9	18.3	87.7	**0.0062**	0.014	223.6	ug/L	17	Standard
	Al	27	17488.4	2.4	**0.0832**	0.003	3.2	ug/L	1410	Standard
	Sc	45	39132.2	2.7				ug/L	39543	Standard
	Ti	47	49.7	51.3	**0.0353**	0.101	286.1	ug/L	39	Standard
	V	51	1251.1	3.7	**-0.0255**	0.008	31.0	ug/L	1415	Standard
	Cr	52	8706.2	0.9	**0.2013**	0.013	6.2	ug/L	6946	Standard
	Cr	53	700.0	11.0	**0.4613**	0.093	20.2	ug/L	328	Standard
	Mn	55	1542.7	1.5	**0.0910**	0.006	6.6	ug/L	823	Standard
	Co	59	278.7	9.1	**0.0085**	0.003	38.3	ug/L	252	Standard
	Ni	60	602.0	5.9	**0.2166**	0.018	8.1	ug/L	192	Standard
	Cu	65	210.7	10.6	**0.0207**	0.012	58.0	ug/L	141	Standard
	Zn	66	1817.8	2.7	**0.8408**	0.028	3.4	ug/L	491	Standard
>	**Ge**	72	511962.6	1.0				ug/L	515087	Standard
	As	75	-35.9	60.1	**0.0082**	0.019	234.7	ug/L	20	Standard
	Se	82	20.1	18.9	**0.0563**	0.034	59.9	ug/L	15	Standard
	Se-1	77	44.7	10.1	**-0.1225**	0.058	47.0	ug/L	48	Standard
>	**Ga**	71	48.3	15.8				mg/L	90	Standard
	Rb	85	66.7	31.2				ug/L	23	Standard
	Y	89	442200.8	1.3				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	193.0	16.8	**0.0509**	0.011	21.8	ug/L	11	Standard
	Ag	107	97.0	17.0	**-0.0021**	0.002	110.1	ug/L	134	Standard
	Cd	111	7.4	64.2	**0.0005**	0.002	466.2	mg/L	7	Standard
	Cd	114	31.4	28.6	**0.0010**	0.002	163.8	ug/L	16	Standard
>	**In**	115	622890.2	1.7				ug/L	611117	Standard
	Sn	118	691.7	5.6	**0.0128**	0.005	38.1	ug/L	518	Standard
	Sb	123	212.1	15.5	**0.0312**	0.006	20.3	ug/L	127	Standard
	Ba	135	298.3	2.4	**0.1167**	0.001	1.2	ug/L	45	Standard
	Ce	140	113.3	24.3				ug/L	298	Standard
>	**Tb**	159	998579.7	1.5				ug/L	977942	Standard
	Ho	165	16.7	17.3				ug/L	12	Standard
	Tl	203	39.0	48.9	**-0.0004**	0.002	526.4	ug/L	31	Standard
	Tl	205	20.0	50.0	**0.0003**	0.001	421.9	ug/L	20	Standard
	Pb	206	313.7	9.3	**-0.0086**	0.005	56.5	ug/L	286	Standard
	Pb	207	273.0	2.4	**-0.0050**	0.001	28.4	ug/L	257	Standard
	Pb	208	1093.3	1.9	**-0.0055**	0.001	16.3	ug/L	956	Standard
	U	238	24.3	55.2	**0.0002**	0.002	711.6	ug/L	17	Standard
>	**Bi**	209	587793.8	0.5				ug/L	557080	Standard

Sample ID: PBS M7 WG573333-03
Report Date/Time: Tuesday, June 28, 2016 14:40:20
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	36.7	7.9	**-0.0078**	0.003	35.7	mg/L	32	Standard
K	39	35.0	51.5	**-0.0208**	0.064	309.0	mg/L	30	Standard
Ca	43	25.0	34.6	**-1.5082**	1.327	88.0	mg/L	48	Standard
Fe	54	165.5	15.5	**0.0140**	0.019	138.8	mg/L	159	Standard
Fe	57	203.3	9.9	**0.0704**	0.084	119.9	mg/L	195	Standard
> **Sc-1**	45	39132.2	2.7				mg/L	39543	Standard
Cl	35	33734.8	0.7				ug/L	27642	Standard
Kr	83	2.3	65.5				ug/L	2	Standard
Br	81	1786.8	3.4				ug/L	1647	Standard
P	31	17722.0	1.9				ug/L	15231	Standard
S	34	2341.8	1.1				ug/L	1845	Standard
Sr	88	65.0	27.7				ug/L	70	Standard
C	12	180.0	47.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	18.7	93.3				mg/L	12	Standard
Ho-1	165	16.7	17.3				mg/L	12	Standard
Er	166	26.7	21.7				mg/L	20	Standard
I	127	1696.8	8.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		116.256	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.393	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBS M7 WG573333-03
Report Date/Time: Tuesday, June 28, 2016 14:40:20
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.927
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.513
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: LCSS M7 WG573333-04
Sample Date/Time: Tuesday, June 28, 2016 14:41:14
Number of Replicates: 3
Autosampler Position: 216
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92927.6	3.1				ug/L	81760	Standard
	Be	9	30111.6	3.7	**26.7031**	1.769	6.6	ug/L	17	Standard
	Al	27	18035.9	13.5	**0.0888**	0.016	18.2	ug/L	1410	Standard
	Sc	45	39361.1	1.9				ug/L	39543	Standard
	Ti	47	37.0	28.2	**-0.0130**	0.042	322.2	ug/L	39	Standard
	V	51	182372.3	3.8	**26.6531**	1.518	5.7	ug/L	1415	Standard
	Cr	52	189442.3	3.6	**26.8012**	1.501	5.6	ug/L	6946	Standard
	Cr	53	22922.3	1.8	**26.3866**	1.122	4.3	ug/L	328	Standard
	Mn	55	169435.1	3.2	**26.6017**	1.285	4.8	ug/L	823	Standard
	Co	59	193981.0	3.8	**28.2331**	1.651	5.8	ug/L	252	Standard
	Ni	60	48862.7	3.3	**26.4525**	1.352	5.1	ug/L	192	Standard
	Cu	65	49743.6	2.7	**26.7036**	1.173	4.4	ug/L	141	Standard
	Zn	66	30362.1	3.5	**27.5261**	1.455	5.3	ug/L	491	Standard
>	**Ge**	72	503563.1	2.3				ug/L	515087	Standard
	As	75	29155.1	3.3	**26.5827**	1.325	5.0	ug/L	20	Standard
	Se	82	3055.7	1.7	**27.0468**	1.107	4.1	ug/L	15	Standard
	Se-1	77	2068.5	3.5	**27.3757**	1.137	4.2	ug/L	48	Standard
>	**Ga**	71	53.3	14.3				mg/L	90	Standard
	Rb	85	78.3	16.1				ug/L	23	Standard
	Y	89	444634.7	2.0				ug/L	448510	Standard
>	**Rh**	103	18.3	41.7				ug/L	15	Standard
	Mo	98	104.5	23.9	**0.0190**	0.010	51.3	ug/L	11	Standard
	Ag	107	179344.1	3.7	**26.2092**	1.706	6.5	ug/L	134	Standard
	Cd	111	55845.5	2.3	**25.6355**	1.334	5.2	mg/L	7	Standard
	Cd	114	136877.4	2.6	**25.8483**	1.371	5.3	ug/L	16	Standard
>	**In**	115	623373.0	3.0				ug/L	611117	Standard
	Sn	118	755.0	11.3	**0.0235**	0.016	68.2	ug/L	518	Standard
	Sb	123	124890.1	3.5	**25.9676**	1.618	6.2	ug/L	127	Standard
	Ba	135	53944.6	3.3	**25.6212**	1.575	6.1	ug/L	45	Standard
	Ce	140	106.7	9.8				ug/L	298	Standard
>	**Tb**	159	976375.9	3.3				ug/L	977942	Standard
	Ho	165	15.0	57.7				ug/L	12	Standard
	Tl	203	229696.9	2.8	**26.5184**	1.549	5.8	ug/L	31	Standard
	Tl	205	205993.8	3.8	**27.3937**	1.852	6.8	ug/L	20	Standard
	Pb	206	150555.8	3.2	**27.0395**	1.637	6.1	ug/L	286	Standard
	Pb	207	130667.9	3.7	**26.0217**	1.699	6.5	ug/L	257	Standard
	Pb	208	532429.0	3.5	**27.4071**	1.756	6.4	ug/L	956	Standard
	U	238	197217.2	3.4	**26.7798**	1.720	6.4	ug/L	17	Standard
>	**Bi**	209	573273.6	3.1				ug/L	557080	Standard

Sample ID: LCSS M7 WG573333-04
Report Date/Time: Tuesday, June 28, 2016 14:43:31
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**21.2655**	18.412	86.6	mg/L	0	Standard
Mg	24	50.0	43.6	**0.0096**	0.028	296.2	mg/L	32	Standard
K	39	30.0	100.0	**-0.0395**	0.101	254.9	mg/L	30	Standard
Ca	43	25.0	40.0	**-1.5307**	1.509	98.6	mg/L	48	Standard
Fe	54	141.7	2.7	**-0.0082**	0.003	33.7	mg/L	159	Standard
Fe	57	226.7	9.9	**0.1435**	0.065	45.0	mg/L	195	Standard
> **Sc-1**	45	39361.1	1.9				mg/L	39543	Standard
Cl	35	35061.2	2.7				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1886.8	4.9				ug/L	1647	Standard
P	31	18956.9	6.3				ug/L	15231	Standard
S	34	2375.2	1.6				ug/L	1845	Standard
Sr	88	65.0	30.8				ug/L	70	Standard
C	12	170.0	38.6				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	12.7	49.8				mg/L	12	Standard
Ho-1	165	15.0	57.7				mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	1553.4	15.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.659	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.763	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSS M7 WG573333-04
Report Date/Time: Tuesday, June 28, 2016 14:43:31
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.006
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.907
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSS M7 WG573333-04
Report Date/Time: Tuesday, June 28, 2016 14:43:31
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606105301 WG573333-02
Sample Date/Time: Tuesday, June 28, 2016 14:44:26
Number of Replicates: 3
Autosampler Position: 217
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92504.9	1.8				ug/L	81760	Standard
	Be	9	211.7	3.6	**0.1787**	0.009	5.2	ug/L	17	Standard
	Al	27	476632.7	0.6	**2.7192**	0.058	2.1	ug/L	1410	Standard
	Sc	45	39899.1	1.7				ug/L	39543	Standard
	Ti	47	7508.9	2.4	**29.9625**	0.867	2.9	ug/L	39	Standard
	V	51	54226.8	0.3	**7.5775**	0.105	1.4	ug/L	1415	Standard
	Cr	52	35510.9	0.6	**4.0327**	0.059	1.5	ug/L	6946	Standard
	Cr	53	3988.9	7.5	**4.1910**	0.323	7.7	ug/L	328	Standard
	Mn	55	226862.1	2.5	**34.7724**	1.274	3.7	ug/L	823	Standard
	Co	59	8803.9	0.2	**1.2188**	0.017	1.4	ug/L	252	Standard
	Ni	60	5989.2	1.6	**3.0677**	0.039	1.3	ug/L	192	Standard
	Cu	65	4442.7	0.7	**2.2418**	0.014	0.6	ug/L	141	Standard
	Zn	66	19698.8	1.0	**17.1048**	0.367	2.1	ug/L	491	Standard
>	**Ge**	72	516312.9	1.2				ug/L	515087	Standard
	As	75	2960.6	0.8	**2.6677**	0.048	1.8	ug/L	20	Standard
	Se	82	53.9	7.9	**0.3477**	0.042	12.1	ug/L	15	Standard
	Se-1	77	82.0	6.8	**0.3676**	0.085	23.1	ug/L	48	Standard
>	**Ga**	71	5075.9	3.2				mg/L	90	Standard
	Rb	85	19981.5	3.8				ug/L	23	Standard
	Y	89	472017.4	2.4				ug/L	448510	Standard
>	**Rh**	103	21.7	26.6				ug/L	15	Standard
	Mo	98	6655.2	1.4	**2.4116**	0.043	1.8	ug/L	11	Standard
	Ag	107	178.7	11.3	**0.0101**	0.003	29.0	ug/L	134	Standard
	Cd	111	244.9	9.1	**0.1102**	0.010	9.1	mg/L	7	Standard
	Cd	114	621.2	5.8	**0.1131**	0.007	6.2	ug/L	16	Standard
>	**In**	115	619055.2	0.3				ug/L	611117	Standard
	Sn	118	640.0	5.5	**0.0049**	0.006	124.7	ug/L	518	Standard
	Sb	123	463.5	5.3	**0.0841**	0.005	6.4	ug/L	127	Standard
	Ba	135	41539.5	1.0	**19.8378**	0.203	1.0	ug/L	45	Standard
	Ce	140	324520.8	1.6				ug/L	298	Standard
>	**Tb**	159	988287.6	1.4				ug/L	977942	Standard
	Ho	165	2263.5	6.3				ug/L	12	Standard
	Tl	203	550.3	1.1	**0.0581**	0.001	1.2	ug/L	31	Standard
	Tl	205	506.7	1.1	**0.0644**	0.001	1.1	ug/L	20	Standard
	Pb	206	143402.5	0.9	**25.4978**	0.319	1.3	ug/L	286	Standard
	Pb	207	120708.6	0.6	**23.7943**	0.166	0.7	ug/L	257	Standard
	Pb	208	491539.2	1.4	**25.0461**	0.389	1.6	ug/L	956	Standard
	U	238	1919.1	2.3	**0.2551**	0.006	2.5	ug/L	17	Standard
>	**Bi**	209	578289.7	0.5				ug/L	557080	Standard

Sample ID: L1606105301 WG573333-02
Report Date/Time: Tuesday, June 28, 2016 14:46:43
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.6918**	18.510	173.1	mg/L	0	Standard
Mg	24	41.7	30.2	**-0.0023**	0.016	678.2	mg/L	32	Standard
K	39	48.3	15.8	**0.0204**	0.026	126.1	mg/L	30	Standard
Ca	43	38.3	27.2	**0.4078**	1.458	357.4	mg/L	48	Standard
Fe	54	3682.2	3.4	**3.1665**	0.134	4.2	mg/L	159	Standard
Fe	57	1100.0	9.6	**3.0601**	0.408	13.3	mg/L	195	Standard
Sc-1	45	39899.1	1.7				mg/L	39543	Standard
Cl	35	32895.7	2.7				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	2153.5	9.2				ug/L	1647	Standard
P	31	17883.9	3.0				ug/L	15231	Standard
S	34	2195.2	9.1				ug/L	1845	Standard
Sr	88	71.7	4.0				ug/L	70	Standard
C	12	220.0	19.8				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	40.0	25.0				mg/L	0	Standard
Dy	164	4107.1	8.5				mg/L	12	Standard
Ho-1	165	2263.5	6.3				mg/L	12	Standard
Er	166	2160.2	2.1				mg/L	20	Standard
I	127	7528.5	2.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		113.142	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		100.238	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606105301 WG573333-02
Report Date/Time: Tuesday, June 28, 2016 14:46:43
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.299
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.807
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606105302S WG573333-07
Sample Date/Time: Tuesday, June 28, 2016 14:47:37
Number of Replicates: 3
Autosampler Position: 218
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	91483.8	2.5				ug/L	81760	Standard
	Be	9	28575.3	3.7	**25.7021**	0.502	2.0	ug/L	17	Standard
	Al	27	472647.5	0.9	**2.7267**	0.044	1.6	ug/L	1410	Standard
	Sc	45	40355.4	2.6				ug/L	39543	Standard
	Ti	47	8788.9	2.8	**34.8074**	0.551	1.6	ug/L	39	Standard
	V	51	216754.8	1.8	**30.6603**	0.748	2.4	ug/L	1415	Standard
	Cr	52	202224.9	0.6	**27.6955**	0.221	0.8	ug/L	6946	Standard
	Cr	53	25461.4	2.5	**28.3619**	0.467	1.6	ug/L	328	Standard
	Mn	55	374661.4	3.0	**57.0462**	1.611	2.8	ug/L	823	Standard
	Co	59	185021.8	2.1	**26.0317**	0.592	2.3	ug/L	252	Standard
	Ni	60	51684.0	1.6	**27.0535**	0.107	0.4	ug/L	192	Standard
	Cu	65	52133.2	1.4	**27.0623**	0.171	0.6	ug/L	141	Standard
	Zn	66	46187.6	0.8	**40.8778**	0.267	0.7	ug/L	491	Standard
>	**Ge**	72	520472.6	1.2				ug/L	515087	Standard
	As	75	27327.3	2.3	**24.0946**	0.465	1.9	ug/L	20	Standard
	Se	82	2628.8	2.4	**22.4788**	0.502	2.2	ug/L	15	Standard
	Se-1	77	1784.1	1.0	**22.7201**	0.347	1.5	ug/L	48	Standard
>	**Ga**	71	5215.9	5.7				mg/L	90	Standard
	Rb	85	16937.8	2.3				ug/L	23	Standard
	Y	89	458647.4	1.3				ug/L	448510	Standard
>	**Rh**	103	30.0	44.1				ug/L	15	Standard
	Mo	98	18905.8	1.4	**6.8731**	0.106	1.5	ug/L	11	Standard
	Ag	107	155307.6	0.6	**22.7819**	0.123	0.5	ug/L	134	Standard
	Cd	111	52492.7	0.7	**24.1959**	0.317	1.3	mg/L	7	Standard
	Cd	114	128079.2	0.4	**24.2856**	0.076	0.3	ug/L	16	Standard
>	**In**	115	620204.7	0.6				ug/L	611117	Standard
	Sn	118	685.0	20.9	**0.0121**	0.023	191.0	ug/L	518	Standard
	Sb	123	65087.0	0.8	**13.5800**	0.099	0.7	ug/L	127	Standard
	Ba	135	93846.7	0.6	**44.7672**	0.375	0.8	ug/L	45	Standard
	Ce	140	331961.2	1.7				ug/L	298	Standard
>	**Tb**	159	981697.8	0.7				ug/L	977942	Standard
	Ho	165	1903.5	2.0				ug/L	12	Standard
	Tl	203	213525.6	1.3	**24.5875**	0.365	1.5	ug/L	31	Standard
	Tl	205	191177.1	1.7	**25.3529**	0.481	1.9	ug/L	20	Standard
	Pb	206	446032.3	1.2	**80.0233**	1.190	1.5	ug/L	286	Standard
	Pb	207	376762.1	1.5	**74.9381**	1.277	1.7	ug/L	257	Standard
	Pb	208	1506170.9	0.7	**77.4336**	0.690	0.9	ug/L	956	Standard
	U	238	183525.9	0.8	**24.8532**	0.261	1.1	ug/L	17	Standard
>	**Bi**	209	574082.0	0.2				ug/L	557080	Standard

Sample ID: L1606105302S WG573333-07
Report Date/Time: Tuesday, June 28, 2016 14:49:54
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	51.7	22.3	**0.0106**	0.017	161.8	mg/L	32	Standard
K	39	33.3	8.7	**-0.0312**	0.007	21.8	mg/L	30	Standard
Ca	43	38.3	15.1	**0.3768**	1.024	271.7	mg/L	48	Standard
Fe	54	2984.3	5.7	**2.5132**	0.224	8.9	mg/L	159	Standard
Fe	57	1001.7	4.5	**2.6914**	0.198	7.3	mg/L	195	Standard
> **Sc-1**	45	40355.4	2.6				mg/L	39543	Standard
Cl	35	31360.4	1.4				ug/L	27642	Standard
Kr	83	1.0	173.2				ug/L	2	Standard
Br	81	2753.6	4.6				ug/L	1647	Standard
P	31	16745.9	1.5				ug/L	15231	Standard
S	34	2118.5	5.5				ug/L	1845	Standard
Sr	88	78.3	9.8				ug/L	70	Standard
C	12	286.7	29.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	56.7	100.3				mg/L	0	Standard
Dy	164	3434.6	3.8				mg/L	12	Standard
Ho-1	165	1903.5	2.0				mg/L	12	Standard
Er	166	1860.1	3.3				mg/L	20	Standard
I	127	5512.7	1.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.893	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.046	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606105302S WG573333-07
Report Date/Time: Tuesday, June 28, 2016 14:49:54
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.487
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.052
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606105303SD WG573333-08
Sample Date/Time: Tuesday, June 28, 2016 14:50:49
Number of Replicates: 3
Autosampler Position: 219
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90707.5	2.1				ug/L	81760	Standard
	Be	9	28490.1	1.5	**25.8507**	0.239	0.9	ug/L	17	Standard
	Al	27	517242.7	0.8	**3.0108**	0.045	1.5	ug/L	1410	Standard
	Sc	45	41104.0	1.1				ug/L	39543	Standard
	Ti	47	9706.8	1.9	**39.3684**	1.018	2.6	ug/L	39	Standard
	V	51	220560.4	0.1	**31.9323**	0.545	1.7	ug/L	1415	Standard
	Cr	52	204805.0	1.1	**28.7389**	0.342	1.2	ug/L	6946	Standard
	Cr	53	25147.5	3.9	**28.6674**	1.000	3.5	ug/L	328	Standard
	Mn	55	371807.0	1.2	**57.9433**	1.578	2.7	ug/L	823	Standard
	Co	59	187737.3	0.6	**27.0316**	0.558	2.1	ug/L	252	Standard
	Ni	60	52302.5	0.6	**28.0202**	0.336	1.2	ug/L	192	Standard
	Cu	65	51530.8	0.8	**27.3738**	0.234	0.9	ug/L	141	Standard
	Zn	66	58955.2	1.6	**53.6399**	0.321	0.6	ug/L	491	Standard
>	**Ge**	72	508670.0	1.6				ug/L	515087	Standard
	As	75	26613.2	0.3	**24.0128**	0.310	1.3	ug/L	20	Standard
	Se	82	2643.5	1.0	**23.1344**	0.377	1.6	ug/L	15	Standard
	Se-1	77	1782.1	1.4	**23.2396**	0.549	2.4	ug/L	48	Standard
>	**Ga**	71	5607.7	2.4				mg/L	90	Standard
	Rb	85	19120.4	0.8				ug/L	23	Standard
	Y	89	464944.5	0.8				ug/L	448510	Standard
>	**Rh**	103	20.0	66.1				ug/L	15	Standard
	Mo	98	6589.1	2.9	**2.3569**	0.065	2.7	ug/L	11	Standard
	Ag	107	161248.5	0.9	**23.3982**	0.239	1.0	ug/L	134	Standard
	Cd	111	52808.4	0.9	**24.0772**	0.233	1.0	mg/L	7	Standard
	Cd	114	128062.2	0.6	**24.0198**	0.192	0.8	ug/L	16	Standard
>	**In**	115	626980.9	0.2				ug/L	611117	Standard
	Sn	118	691.7	6.1	**0.0121**	0.007	57.7	ug/L	518	Standard
	Sb	123	64415.1	1.9	**13.2945**	0.276	2.1	ug/L	127	Standard
	Ba	135	100181.1	0.9	**47.2728**	0.453	1.0	ug/L	45	Standard
	Ce	140	306983.9	0.9				ug/L	298	Standard
>	**Tb**	159	992773.6	1.4				ug/L	977942	Standard
	Ho	165	1925.1	2.4				ug/L	12	Standard
	Tl	203	213931.4	0.2	**24.2345**	0.027	0.1	ug/L	31	Standard
	Tl	205	188875.1	1.2	**24.6411**	0.299	1.2	ug/L	20	Standard
	Pb	206	399455.5	0.3	**70.4964**	0.248	0.4	ug/L	286	Standard
	Pb	207	335397.3	0.3	**65.6206**	0.221	0.3	ug/L	257	Standard
	Pb	208	1377394.4	0.8	**69.6587**	0.536	0.8	ug/L	956	Standard
	U	238	183479.9	0.5	**24.4438**	0.133	0.5	ug/L	17	Standard
>	**Bi**	209	583538.4	0.1				ug/L	557080	Standard

Sample ID: L1606105303SD WG573333-08
Report Date/Time: Tuesday, June 28, 2016 14:53:06
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	45.0	11.1	**0.0005**	0.006	1275.5	mg/L	32	Standard
K	39	46.7	12.4	**0.0102**	0.018	179.3	mg/L	30	Standard
Ca	43	46.7	40.6	**1.4697**	2.764	188.0	mg/L	48	Standard
Fe	54	3552.4	2.1	**2.9565**	0.098	3.3	mg/L	159	Standard
Fe	57	1123.4	10.8	**3.0231**	0.384	12.7	mg/L	195	Standard
> **Sc-1**	45	41104.0	1.1				mg/L	39543	Standard
Cl	35	31493.4	1.6				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	2126.8	11.6				ug/L	1647	Standard
P	31	16392.2	2.4				ug/L	15231	Standard
S	34	1988.5	3.7				ug/L	1845	Standard
Sr	88	56.7	20.4				ug/L	70	Standard
C	12	256.7	29.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	73.3	34.3				mg/L	0	Standard
Dy	164	3598.5	4.6				mg/L	12	Standard
Ho-1	165	1925.1	2.4				mg/L	12	Standard
Er	166	1710.1	5.6				mg/L	20	Standard
I	127	5772.8	6.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.943	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.754	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606105303SD WG573333-08
Report Date/Time: Tuesday, June 28, 2016 14:53:06
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.596
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.749
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606104801
Sample Date/Time: Tuesday, June 28, 2016 14:54:00
Number of Replicates: 3
Autosampler Position: 220
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	87619.5	3.9				ug/L	81760	Standard
	Be	9	323.3	11.2	**0.2951**	0.044	15.0	ug/L	17	Standard
	Al	27	1546436.6	2.8	**9.3502**	0.142	1.5	ug/L	1410	Standard
	Sc	45	40666.2	2.1				ug/L	39543	Standard
	Ti	47	348.0	4.5	**1.2425**	0.055	4.4	ug/L	39	Standard
	V	51	27678.8	0.4	**3.7995**	0.033	0.9	ug/L	1415	Standard
	Cr	52	26992.7	1.5	**2.8437**	0.051	1.8	ug/L	6946	Standard
	Cr	53	2932.0	5.7	**3.0177**	0.170	5.6	ug/L	328	Standard
	Mn	55	139361.0	0.9	**21.4761**	0.101	0.5	ug/L	823	Standard
	Co	59	8883.9	1.8	**1.2405**	0.015	1.2	ug/L	252	Standard
	Ni	60	10083.0	1.9	**5.2809**	0.102	1.9	ug/L	192	Standard
	Cu	65	17655.3	1.5	**9.2565**	0.148	1.6	ug/L	141	Standard
	Zn	66	24172.6	2.4	**21.3574**	0.493	2.3	ug/L	491	Standard
>	Ge	72	512038.2	0.7				ug/L	515087	Standard
	As	75	255.5	6.8	**0.2688**	0.014	5.3	ug/L	20	Standard
	Se	82	53.2	11.2	**0.3458**	0.053	15.2	ug/L	15	Standard
	Se-1	77	80.3	12.8	**0.3545**	0.144	40.6	ug/L	48	Standard
>	Ga	71	3018.6	3.1				mg/L	90	Standard
	Rb	85	16969.5	1.9				ug/L	23	Standard
	Y	89	529226.2	0.9				ug/L	448510	Standard
>	Rh	103	10.0					ug/L	15	Standard
	Mo	98	313.0	4.7	**0.0960**	0.005	5.5	ug/L	11	Standard
	Ag	107	442.0	1.6	**0.0492**	0.001	1.7	ug/L	134	Standard
	Cd	111	534.5	3.9	**0.2456**	0.011	4.4	mg/L	7	Standard
	Cd	114	1438.1	5.6	**0.2703**	0.019	7.1	ug/L	16	Standard
>	In	115	614980.8	1.4				ug/L	611117	Standard
	Sn	118	1005.0	9.0	**0.0670**	0.016	23.2	ug/L	518	Standard
	Sb	123	831.9	19.2	**0.1627**	0.036	21.9	ug/L	127	Standard
	Ba	135	51902.4	2.2	**24.9562**	0.305	1.2	ug/L	45	Standard
	Ce	140	320029.2	2.5				ug/L	298	Standard
>	Tb	159	977861.1	0.7				ug/L	977942	Standard
	Ho	165	6373.0	4.5				ug/L	12	Standard
	Tl	203	276.7	7.5	**0.0277**	0.002	8.0	ug/L	31	Standard
	Tl	205	265.0	17.0	**0.0335**	0.006	17.4	ug/L	20	Standard
	Pb	206	33506.0	0.9	**6.0646**	0.072	1.2	ug/L	286	Standard
	Pb	207	27548.0	1.0	**5.5276**	0.068	1.2	ug/L	257	Standard
	Pb	208	115638.2	1.3	**6.0000**	0.078	1.3	ug/L	956	Standard
	U	238	6967.9	1.8	**0.9584**	0.022	2.3	ug/L	17	Standard
>	Bi	209	563593.3	0.7				ug/L	557080	Standard

Sample ID: L1606104801
Report Date/Time: Tuesday, June 28, 2016 14:56:17
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	48.3	36.3	**0.0057**	0.024	420.3	mg/L	32	Standard
K	39	71.7	21.3	**0.0944**	0.052	54.8	mg/L	30	Standard
Ca	43	58.3	17.8	**3.2603**	1.450	44.5	mg/L	48	Standard
Fe	54	2389.6	6.6	**1.9680**	0.178	9.0	mg/L	159	Standard
Fe	57	746.7	5.6	**1.8289**	0.182	10.0	mg/L	195	Standard
Sc-1	45	40666.2	2.1				mg/L	39543	Standard
Cl	35	32405.3	1.2				ug/L	27642	Standard
Kr	83	4.0	43.3				ug/L	2	Standard
Br	81	2003.5	7.8				ug/L	1647	Standard
P	31	17084.6	3.9				ug/L	15231	Standard
S	34	2108.5	4.8				ug/L	1845	Standard
Sr	88	73.3	30.7				ug/L	70	Standard
C	12	240.0	38.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	20.0				mg/L	0	Standard
Dy	164	9760.8	1.5				mg/L	12	Standard
Ho-1	165	6373.0	4.5				mg/L	12	Standard
Er	166	5897.8	8.3				mg/L	20	Standard
I	127	2141.8	10.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		107.167	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.408	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606104801
Report Date/Time: Tuesday, June 28, 2016 14:56:17
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.632
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.169
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606104801PS WG573475-03
Sample Date/Time: Tuesday, June 28, 2016 14:57:11
Number of Replicates: 3
Autosampler Position: 221
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90811.3	0.7				ug/L	81760	Standard
	Be	9	56498.5	1.8	**51.2132**	1.080	2.1	ug/L	17	Standard
	Al	27	1545775.3	0.6	**9.0143**	0.017	0.2	ug/L	1410	Standard
	Sc	45	41961.4	1.6				ug/L	39543	Standard
	Ti	47	342.0	11.2	**1.1864**	0.157	13.2	ug/L	39	Standard
	V	51	392129.2	0.5	**55.2170**	0.488	0.9	ug/L	1415	Standard
	Cr	52	384433.3	0.4	**53.1980**	0.445	0.8	ug/L	6946	Standard
	Cr	53	48853.7	2.8	**54.3367**	1.767	3.3	ug/L	328	Standard
	Mn	55	478138.1	0.8	**72.3045**	0.492	0.7	ug/L	823	Standard
	Co	59	389747.9	0.3	**54.4630**	0.290	0.5	ug/L	252	Standard
	Ni	60	106262.0	0.3	**55.3242**	0.400	0.7	ug/L	192	Standard
	Cu	65	114943.7	0.4	**59.3369**	0.368	0.6	ug/L	141	Standard
	Zn	66	80976.1	0.1	**71.7501**	0.396	0.6	ug/L	491	Standard
>	**Ge**	72	524333.4	0.5				ug/L	515087	Standard
	As	75	59243.0	1.9	**51.8077**	1.242	2.4	ug/L	20	Standard
	Se	82	6211.1	2.1	**52.8798**	1.213	2.3	ug/L	15	Standard
	Se-1	77	4085.6	1.2	**52.5558**	0.559	1.1	ug/L	48	Standard
>	**Ga**	71	3287.0	1.8				mg/L	90	Standard
	Rb	85	16744.3	3.7				ug/L	23	Standard
	Y	89	522998.5	1.8				ug/L	448510	Standard
>	**Rh**	103	28.3	36.7				ug/L	15	Standard
	Mo	98	332.0	9.1	**0.1017**	0.011	11.1	ug/L	11	Standard
	Ag	107	350941.4	1.0	**51.3750**	0.335	0.7	ug/L	134	Standard
	Cd	111	110107.7	1.0	**50.6313**	0.348	0.7	mg/L	7	Standard
	Cd	114	268982.1	1.5	**50.8849**	0.677	1.3	ug/L	16	Standard
>	**In**	115	621693.0	0.4				ug/L	611117	Standard
	Sn	118	1258.4	5.6	**0.1073**	0.012	10.9	ug/L	518	Standard
	Sb	123	241746.4	1.3	**50.3512**	0.498	1.0	ug/L	127	Standard
	Ba	135	155747.4	0.9	**74.1325**	0.655	0.9	ug/L	45	Standard
	Ce	140	312396.1	0.8				ug/L	298	Standard
>	**Tb**	159	997317.5	1.6				ug/L	977942	Standard
	Ho	165	6548.1	1.9				ug/L	12	Standard
	Tl	203	447292.7	0.3	**50.4657**	0.615	1.2	ug/L	31	Standard
	Tl	205	398193.6	1.5	**51.7351**	0.805	1.6	ug/L	20	Standard
	Pb	206	325033.3	0.1	**57.1116**	0.488	0.9	ug/L	286	Standard
	Pb	207	282464.7	0.8	**55.0228**	0.100	0.2	ug/L	257	Standard
	Pb	208	1144735.7	1.0	**57.6431**	0.938	1.6	ug/L	956	Standard
	U	238	386201.4	1.1	**51.2415**	0.797	1.6	ug/L	17	Standard
>	**Bi**	209	586004.9	0.9				ug/L	557080	Standard

Sample ID: L1606104801PS WG573475-03
Report Date/Time: Tuesday, June 28, 2016 14:59:28
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	48.3	41.8	**0.0038**	0.026	698.9	mg/L	32	Standard
K	39	65.0	20.4	**0.0660**	0.045	68.1	mg/L	30	Standard
Ca	43	46.7	27.0	**1.3158**	1.694	128.8	mg/L	48	Standard
Fe	54	2337.7	11.4	**1.8584**	0.246	13.3	mg/L	159	Standard
Fe	57	760.0	13.2	**1.7916**	0.300	16.7	mg/L	195	Standard
> **Sc-1**	45	41961.4	1.6				mg/L	39543	Standard
Cl	35	32737.3	1.7				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	2180.2	6.6				ug/L	1647	Standard
P	31	17061.3	3.2				ug/L	15231	Standard
S	34	2245.2	3.3				ug/L	1845	Standard
Sr	88	81.7	12.7				ug/L	70	Standard
C	12	226.7	6.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	46.7	32.7				mg/L	0	Standard
Dy	164	10105.0	3.1				mg/L	12	Standard
Ho-1	165	6548.1	1.9				mg/L	12	Standard
Er	166	5884.5	3.8				mg/L	20	Standard
I	127	2158.5	15.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		111.071	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		101.795	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606104801PS WG573475-03
Report Date/Time: Tuesday, June 28, 2016 14:59:28
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.731
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.192
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606104801SDL WG573475-04
Sample Date/Time: Tuesday, June 28, 2016 15:00:23
Number of Replicates: 3
Autosampler Position: 222
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89188.5	2.4				ug/L	81760	Standard
	Be	9	71.7	4.0	**0.0564**	0.001	2.3	ug/L	17	Standard
	Al	27	302683.6	0.7	**1.7866**	0.056	3.1	ug/L	1410	Standard
	Sc	45	41125.8	3.4				ug/L	39543	Standard
	Ti	47	82.7	6.2	**0.1644**	0.030	18.2	ug/L	39	Standard
	V	51	6527.8	1.5	**0.7233**	0.030	4.1	ug/L	1415	Standard
	Cr	52	10512.6	0.6	**0.4384**	0.054	12.4	ug/L	6946	Standard
	Cr	53	805.0	22.0	**0.5666**	0.199	35.2	ug/L	328	Standard
	Mn	55	27749.4	1.2	**4.0892**	0.112	2.7	ug/L	823	Standard
	Co	59	1936.1	6.5	**0.2412**	0.012	4.8	ug/L	252	Standard
	Ni	60	2147.2	1.2	**1.0241**	0.047	4.6	ug/L	192	Standard
	Cu	65	3398.4	1.3	**1.6803**	0.070	4.1	ug/L	141	Standard
	Zn	66	7011.6	1.2	**5.5054**	0.150	2.7	ug/L	491	Standard
>	**Ge**	72	520536.2	3.0				ug/L	515087	Standard
	As	75	60.9	66.9	**0.0932**	0.035	37.4	ug/L	20	Standard
	Se	82	24.2	49.2	**0.0869**	0.099	113.8	ug/L	15	Standard
	Se-1	77	55.3	11.8	**0.0096**	0.103	1081.5	ug/L	48	Standard
>	**Ga**	71	663.3	7.6				mg/L	90	Standard
	Rb	85	3270.4	5.2				ug/L	23	Standard
	Y	89	464717.5	1.9				ug/L	448510	Standard
>	**Rh**	103	16.7	121.2				ug/L	15	Standard
	Mo	98	80.8	3.1	**0.0096**	0.001	7.0	ug/L	11	Standard
	Ag	107	180.0	20.5	**0.0094**	0.005	54.7	ug/L	134	Standard
	Cd	111	113.9	5.5	**0.0481**	0.003	5.7	mg/L	7	Standard
	Cd	114	303.5	12.8	**0.0510**	0.007	13.8	ug/L	16	Standard
>	**In**	115	638226.3	1.3				ug/L	611117	Standard
	Sn	118	801.7	13.2	**0.0279**	0.017	59.8	ug/L	518	Standard
	Sb	123	2112.7	26.2	**0.4158**	0.111	26.6	ug/L	127	Standard
	Ba	135	10186.4	1.4	**4.7001**	0.102	2.2	ug/L	45	Standard
	Ce	140	62112.1	0.6				ug/L	298	Standard
>	**Tb**	159	1019737.0	1.0				ug/L	977942	Standard
	Ho	165	1263.4	4.0				ug/L	12	Standard
	Tl	203	113.0	16.9	**0.0078**	0.002	24.0	ug/L	31	Standard
	Tl	205	118.3	30.6	**0.0128**	0.004	33.6	ug/L	20	Standard
	Pb	206	6901.2	1.0	**1.1348**	0.029	2.6	ug/L	286	Standard
	Pb	207	5663.1	0.8	**1.0321**	0.019	1.9	ug/L	257	Standard
	Pb	208	22745.7	0.6	**1.0713**	0.025	2.3	ug/L	956	Standard
	U	238	1347.1	2.4	**0.1734**	0.003	1.5	ug/L	17	Standard
>	**Bi**	209	593755.8	1.9				ug/L	557080	Standard

Sample ID: L1606104801SDL WG573475-04
Report Date/Time: Tuesday, June 28, 2016 15:02:40
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	45.0	40.1	**0.0009**	0.025	2626.8	mg/L	32	Standard
K	39	30.0	28.9	**-0.0443**	0.026	58.3	mg/L	30	Standard
Ca	43	28.3	20.4	**-1.2169**	0.752	61.8	mg/L	48	Standard
Fe	54	584.0	13.3	**0.3714**	0.068	18.2	mg/L	159	Standard
Fe	57	295.0	14.5	**0.3314**	0.118	35.5	mg/L	195	Standard
> **Sc-1**	45	41125.8	3.4				mg/L	39543	Standard
Cl	35	31826.1	1.1				ug/L	27642	Standard
Kr	83	2.7	21.7				ug/L	2	Standard
Br	81	1903.5	3.4				ug/L	1647	Standard
P	31	16692.5	3.1				ug/L	15231	Standard
S	34	2128.5	4.6				ug/L	1845	Standard
Sr	88	68.3	36.1				ug/L	70	Standard
C	12	176.7	42.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	1953.3	7.3				mg/L	12	Standard
Ho-1	165	1263.4	4.0				mg/L	12	Standard
Er	166	1116.7	6.8				mg/L	20	Standard
I	127	1450.1	11.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		109.086	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.058	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606104801SDL WG573475-04
Report Date/Time: Tuesday, June 28, 2016 15:02:40
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.436
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.584
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Tuesday, June 28, 2016 15:03:36
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	89282.3	1.4				ug/L	81760	Standard
	Be	9	15.0	33.3	**0.0041**	0.005	114.8	ug/L	17	Standard
	Al	27	7897505.8	1.4	**46.9037**	0.018	0.0	ug/L	1410	Standard
	Sc	45	39757.1	3.6				ug/L	39543	Standard
	Ti	47	27232.5	2.5	**109.9775**	3.234	2.9	ug/L	39	Standard
	V	51	789.3	35.9	**-0.0922**	0.042	45.6	ug/L	1415	Standard
	Cr	52	7275.7	2.2	**-0.0059**	0.028	478.1	ug/L	6946	Standard
	Cr	53	6823.2	10.4	**7.4758**	0.819	11.0	ug/L	328	Standard
	Mn	55	1397.7	2.4	**0.0683**	0.004	5.9	ug/L	823	Standard
	Co	59	388.0	3.6	**0.0242**	0.002	7.5	ug/L	252	Standard
	Ni	60	839.0	6.3	**0.3431**	0.028	8.3	ug/L	192	Standard
	Cu	65	614.0	0.9	**0.2341**	0.001	0.3	ug/L	141	Standard
	Zn	66	1957.1	1.8	**0.9682**	0.037	3.8	ug/L	491	Standard
>	Ge	72	512179.1	1.0				ug/L	515087	Standard
	As	75	-64.0	70.1	**-0.0170**	0.040	237.0	ug/L	20	Standard
	Se	82	21.9	13.2	**0.0720**	0.027	37.3	ug/L	15	Standard
	Se-1	77	397.7	7.2	**4.5923**	0.428	9.3	ug/L	48	Standard
>	Ga	71	168.3	7.5				mg/L	90	Standard
	Rb	85	1135.0	8.9				ug/L	23	Standard
	Y	89	442720.4	1.1				ug/L	448510	Standard
>	Rh	103	10.0	50.0				ug/L	15	Standard
	Mo	98	272741.3	0.7	**100.8481**	1.185	1.2	ug/L	11	Standard
	Ag	107	262.0	2.0	**0.0228**	0.001	4.1	ug/L	134	Standard
	Cd	111	-198.2	9.3	**-0.0956**	0.008	8.7	mg/L	7	Standard
	Cd	114	465.1	17.4	**0.0845**	0.015	17.7	ug/L	16	Standard
>	In	115	611378.2	0.9				ug/L	611117	Standard
	Sn	118	743.4	13.4	**0.0238**	0.017	73.3	ug/L	518	Standard
	Sb	123	656.7	25.2	**0.1264**	0.036	28.5	ug/L	127	Standard
	Ba	135	610.7	6.5	**0.2705**	0.017	6.2	ug/L	45	Standard
	Ce	140	266.7	9.2				ug/L	298	Standard
>	Tb	159	977869.4	2.5				ug/L	977942	Standard
	Ho	165	16.7	17.3				ug/L	12	Standard
	Tl	203	193.3	8.8	**0.0182**	0.002	10.2	ug/L	31	Standard
	Tl	205	168.3	10.4	**0.0207**	0.002	10.8	ug/L	20	Standard
	Pb	206	552.7	5.7	**0.0387**	0.005	12.2	ug/L	286	Standard
	Pb	207	474.0	2.6	**0.0392**	0.001	3.3	ug/L	257	Standard
	Pb	208	1934.7	0.9	**0.0422**	0.002	5.1	ug/L	956	Standard
	U	238	35.3	29.7	**0.0020**	0.001	71.8	ug/L	17	Standard
>	Bi	209	556811.8	1.3				ug/L	557080	Standard

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 15:05:53
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**20.5038**	35.505	173.2	mg/L	0	Standard
Mg	24	10406.9	1.4	**13.7318**	0.472	3.4	mg/L	32	Standard
K	39	1353.4	4.9	**4.4093**	0.243	5.5	mg/L	30	Standard
Ca	43	183.3	3.1	**22.3899**	1.779	7.9	mg/L	48	Standard
Fe	54	14710.0	1.7	**13.1205**	0.693	5.3	mg/L	159	Standard
Fe	57	4182.2	3.4	**13.4196**	0.181	1.3	mg/L	195	Standard
Sc-1	45	39757.1	3.6				mg/L	39543	Standard
Cl	35	33087.4	2.4				ug/L	27642	Standard
Kr	83	1.3	173.2				ug/L	2	Standard
Br	81	1766.8	6.4				ug/L	1647	Standard
P	31	17728.7	3.7				ug/L	15231	Standard
S	34	2340.2	3.3				ug/L	1845	Standard
Sr	88	45.0	0.0				ug/L	70	Standard
C	12	306.7	6.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	93.3	12.4				mg/L	0	Standard
Dy	164	15.1	45.8				mg/L	12	Standard
Ho-1	165	16.7	17.3				mg/L	12	Standard
Er	166	33.3	75.5				mg/L	20	Standard
I	127	1891.8	6.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27	0.938		
Sc	45			
Ti	47	109.978		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.435	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 15:05:53
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	100.848		
Ag	107			
Cd	111			
Cd	114			
> In	115		100.043	
Sn	118			
Sb	123			
Ba	135			
Ce	140			
> Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
> Bi	209		99.952	
Na	23	164.030		
Mg	24	274.635		
K	39	88.187		
Ca	43	149.266		
Fe	54	104.964		
Fe	57	107.357		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Na	23	
QC Std 4	Mg	24	

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 15:05:53
Page 3

Approved: June 29, 2016

QC Std 4	Ca	43

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 15:05:53
Page 4

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Tuesday, June 28, 2016 15:06:47
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87968.0	1.5				ug/L	81760	Standard
	Be	9	111223.4	1.1	**104.1090**	2.678	2.6	ug/L	17	Standard
	Al	27	8303576.6	1.0	**50.0632**	1.083	2.2	ug/L	1410	Standard
	Sc	45	38970.1	2.5				ug/L	39543	Standard
	Ti	47	28030.6	1.5	**112.6949**	1.093	1.0	ug/L	39	Standard
	V	51	696349.8	1.4	**100.1131**	1.784	1.8	ug/L	1415	Standard
	Cr	52	683707.3	0.8	**97.2904**	0.687	0.7	ug/L	6946	Standard
	Cr	53	91713.5	1.8	**104.2725**	1.761	1.7	ug/L	328	Standard
	Mn	55	633248.9	0.6	**97.6578**	0.701	0.7	ug/L	823	Standard
	Co	59	716430.9	0.3	**102.0719**	1.125	1.1	ug/L	252	Standard
	Ni	60	180604.4	0.8	**95.9232**	1.625	1.7	ug/L	192	Standard
	Cu	65	185052.2	0.7	**97.4326**	1.562	1.6	ug/L	141	Standard
	Zn	66	113502.2	0.3	**102.8618**	0.650	0.6	ug/L	491	Standard
>	**Ge**	72	514438.0	0.9				ug/L	515087	Standard
	As	75	114188.1	0.3	**101.7361**	0.865	0.8	ug/L	20	Standard
	Se	82	11752.5	0.4	**102.0988**	1.402	1.4	ug/L	15	Standard
	Se-1	77	8079.2	0.7	**106.6610**	0.727	0.7	ug/L	48	Standard
>	**Ga**	71	190.0	7.0				mg/L	90	Standard
	Rb	85	893.4	6.0				ug/L	23	Standard
	Y	89	434662.6	1.6				ug/L	448510	Standard
>	**Rh**	103	38.3	52.7				ug/L	15	Standard
	Mo	98	282406.5	1.5	**107.3339**	1.537	1.4	ug/L	11	Standard
	Ag	107	594345.7	0.8	**90.9666**	1.393	1.5	ug/L	134	Standard
	Cd	111	207638.0	0.2	**99.8149**	1.587	1.6	mg/L	7	Standard
	Cd	114	499764.2	0.7	**98.8323**	1.071	1.1	ug/L	16	Standard
>	**In**	115	594810.8	1.6				ug/L	611117	Standard
	Sn	118	1138.4	7.8	**0.0959**	0.015	15.6	ug/L	518	Standard
	Sb	123	463528.1	1.1	**100.9493**	2.612	2.6	ug/L	127	Standard
	Ba	135	197524.5	0.1	**98.2892**	1.431	1.5	ug/L	45	Standard
	Ce	140	85.0	21.2				ug/L	298	Standard
>	**Tb**	159	967938.8	1.8				ug/L	977942	Standard
	Ho	165	8.3	69.3				ug/L	12	Standard
	Tl	203	848553.4	0.8	**99.1555**	1.255	1.3	ug/L	31	Standard
	Tl	205	762572.5	0.3	**102.6175**	1.400	1.4	ug/L	20	Standard
	Pb	206	559247.5	0.6	**101.8213**	1.258	1.2	ug/L	286	Standard
	Pb	207	488004.9	0.6	**98.5027**	1.226	1.2	ug/L	257	Standard
	Pb	208	1955600.4	0.4	**102.0354**	1.641	1.6	ug/L	956	Standard
	U	238	737824.6	0.7	**101.3974**	2.083	2.1	ug/L	17	Standard
>	**Bi**	209	565855.3	1.5				ug/L	557080	Standard

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 15:09:04
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**21.8863**	18.964	86.6	mg/L	0	Standard
Mg	24	10742.1	1.4	**14.4601**	0.490	3.4	mg/L	32	Standard
K	39	1261.7	7.3	**4.1815**	0.232	5.5	mg/L	30	Standard
Ca	43	183.3	29.9	**22.9918**	8.781	38.2	mg/L	48	Standard
Fe	54	15164.6	1.1	**13.7950**	0.409	3.0	mg/L	159	Standard
Fe	57	4003.9	4.1	**13.0952**	0.606	4.6	mg/L	195	Standard
Sc-1	45	38970.1	2.5				mg/L	39543	Standard
Cl	35	31476.7	2.0				ug/L	27642	Standard
Kr	83	3.3	75.5				ug/L	2	Standard
Br	81	1726.8	7.1				ug/L	1647	Standard
P	31	16804.3	1.0				ug/L	15231	Standard
S	34	2221.8	2.4				ug/L	1845	Standard
Sr	88	46.7	22.3				ug/L	70	Standard
C	12	280.0	21.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	0	Standard
Dy	164	25.2	45.9				mg/L	12	Standard
Ho-1	165	8.3	69.3				mg/L	12	Standard
Er	166	30.0	66.7				mg/L	20	Standard
I	127	1238.4	19.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	104.109		
Al	27	1.001		
Sc	45			
Ti	47	112.695		
V	51	100.113		
Cr	52	97.290		
Cr	53			
Mn	55	97.658		
Co	59	102.072		
Ni	60	95.923		
Cu	65	97.433		
Zn	66	102.862		
Ge	72		99.874	
As	75	101.736		
Se	82	102.099		
Se-1	77			
Ga	71			

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 15:09:04
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	107.334		
Ag	107	90.967		
Cd	111	99.815		
Cd	114			
In	115		97.332	
Sn	118			
Sb	123	100.949		
Ba	135	98.289		
Ce	140			
Tb	159			
Ho	165			
Tl	203	99.155		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.035		
U	238	101.397		
Bi	209		101.575	
Na	23	175.090		
Mg	24	289.203		
K	39	83.630		
Ca	43	153.278		
Fe	54	110.360		
Fe	57	104.761		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Na	23	
QC Std 5	Mg	24	

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 15:09:04
Page 3

Approved: June 29, 2016

QC Std 5	Ca	43

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 15:09:04
Page 4

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 15:10:01
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85172.2	3.3				ug/L	81760	Standard
	Be	9	53236.8	4.1	**51.4767**	2.435	4.7	ug/L	17	Standard
	Al	27	7628795.8	0.6	**47.5256**	1.416	3.0	ug/L	1410	Standard
	Sc	45	37696.8	2.0				ug/L	39543	Standard
	Ti	47	24164.9	0.9	**98.0401**	2.312	2.4	ug/L	39	Standard
	V	51	343835.2	0.1	**49.7831**	0.751	1.5	ug/L	1415	Standard
	Cr	52	344282.1	1.0	**48.9253**	1.054	2.2	ug/L	6946	Standard
	Cr	53	43472.3	1.9	**49.7131**	1.693	3.4	ug/L	328	Standard
	Mn	55	313567.9	1.4	**48.7360**	1.405	2.9	ug/L	823	Standard
	Co	59	359786.6	1.1	**51.7171**	1.082	2.1	ug/L	252	Standard
	Ni	60	90351.3	0.3	**48.3759**	0.827	1.7	ug/L	192	Standard
	Cu	65	92251.0	0.7	**48.9699**	0.708	1.4	ug/L	141	Standard
	Zn	66	54515.2	1.1	**49.4372**	1.191	2.4	ug/L	491	Standard
>	**Ge**	72	509796.1	1.5				ug/L	515087	Standard
	As	75	55417.4	0.8	**49.8530**	1.118	2.2	ug/L	20	Standard
	Se	82	5885.8	2.0	**51.5527**	1.791	3.5	ug/L	15	Standard
	Se-1	77	3913.8	2.3	**51.7920**	1.979	3.8	ug/L	48	Standard
>	**Ga**	71	60.0	14.4				mg/L	90	Standard
	Rb	85	760.0	3.9				ug/L	23	Standard
	Y	89	437496.8	1.0				ug/L	448510	Standard
>	**Rh**	103	35.0	28.6				ug/L	15	Standard
	Mo	98	271668.8	0.6	**101.2993**	0.747	0.7	ug/L	11	Standard
	Ag	107	329448.9	0.2	**49.4598**	0.339	0.7	ug/L	134	Standard
	Cd	111	103364.2	1.7	**48.7409**	0.531	1.1	mg/L	7	Standard
	Cd	114	251263.5	1.7	**48.7432**	0.553	1.1	ug/L	16	Standard
>	**In**	115	606236.0	0.6				ug/L	611117	Standard
	Sn	118	288784.4	0.3	**49.1320**	0.341	0.7	ug/L	518	Standard
	Sb	123	235993.4	0.4	**50.4089**	0.456	0.9	ug/L	127	Standard
	Ba	135	99324.5	0.4	**48.4735**	0.237	0.5	ug/L	45	Standard
	Ce	140	126.7	35.6				ug/L	298	Standard
>	**Tb**	159	959991.8	0.1				ug/L	977942	Standard
	Ho	165	8.3	34.6				ug/L	12	Standard
	Tl	203	427212.2	0.6	**50.5774**	0.492	1.0	ug/L	31	Standard
	Tl	205	380951.0	2.1	**51.9422**	1.422	2.7	ug/L	20	Standard
	Pb	206	273119.7	0.4	**50.3511**	0.572	1.1	ug/L	286	Standard
	Pb	207	248886.2	0.4	**50.8720**	0.472	0.9	ug/L	257	Standard
	Pb	208	972505.1	1.3	**51.3811**	0.993	1.9	ug/L	956	Standard
	U	238	367751.0	1.4	**51.2030**	1.055	2.1	ug/L	17	Standard
>	**Bi**	209	558440.3	0.7				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:12:17
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4133.9	5.5	**5.7164**	0.326	5.7	mg/L	32	Standard
K	39	1253.4	8.2	**4.3011**	0.352	8.2	mg/L	30	Standard
Ca	43	70.0	37.8	**5.7557**	4.052	70.4	mg/L	48	Standard
Fe	54	5742.0	2.6	**5.3173**	0.253	4.8	mg/L	159	Standard
Fe	57	1578.4	3.1	**4.9652**	0.143	2.9	mg/L	195	Standard
Sc-1	45	37696.8	2.0				mg/L	39543	Standard
Cl	35	31580.2	0.9				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	1660.1	9.6				ug/L	1647	Standard
P	31	16826.0	2.3				ug/L	15231	Standard
S	34	2313.5	5.0				ug/L	1845	Standard
Sr	88	61.7	26.1				ug/L	70	Standard
C	12	133.3	18.9				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	13.3	114.6				mg/L	12	Standard
Ho-1	165	8.3	34.6				mg/L	12	Standard
Er	166	0.0					mg/L	20	Standard
I	127	1290.1	24.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	102.953		
Al	27	95.051		
Sc	45			
Ti	47	98.040		
V	51	99.566		
Cr	52	97.851		
Cr	53			
Mn	55	97.472		
Co	59	103.434		
Ni	60	96.752		
Cu	65	97.940		
Zn	66	98.874		
Ge	72		98.973	
As	75	99.706		
Se	82	103.105		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:12:17
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.299		
Ag	107	98.920		
Cd	111	97.482		
Cd	114			
In	115			99.201
Sn	118	98.264		
Sb	123	100.818		
Ba	135	96.947		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.155		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.762		
U	238	102.406		
Bi	209			100.244
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:12:17
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 15:13:12
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88670.5	2.8				ug/L	81760	Standard
	Be	9	51.7	24.4	**0.0381**	0.010	27.4	ug/L	17	Standard
	Al	27	5021.0	57.4	**0.0153**	0.016	105.4	ug/L	1410	Standard
	Sc	45	39924.2	3.2				ug/L	39543	Standard
	Ti	47	36.7	33.9	**-0.0173**	0.048	276.4	ug/L	39	Standard
	V	51	1293.4	14.6	**-0.0198**	0.025	126.5	ug/L	1415	Standard
	Cr	52	6153.2	0.9	**-0.1686**	0.020	12.1	ug/L	6946	Standard
	Cr	53	480.0	7.9	**0.2078**	0.038	18.2	ug/L	328	Standard
	Mn	55	818.0	20.7	**-0.0219**	0.025	112.7	ug/L	823	Standard
	Co	59	362.0	35.7	**0.0203**	0.018	87.2	ug/L	252	Standard
	Ni	60	216.7	22.5	**0.0105**	0.024	233.6	ug/L	192	Standard
	Cu	65	206.0	48.8	**0.0177**	0.051	290.8	ug/L	141	Standard
	Zn	66	512.3	23.8	**-0.3586**	0.106	29.6	ug/L	491	Standard
>	**Ge**	72	512537.5	1.6				ug/L	515087	Standard
	As	75	16.6	529.1	**0.0544**	0.078	142.5	ug/L	20	Standard
	Se	82	20.8	66.0	**0.0610**	0.117	191.3	ug/L	15	Standard
	Se-1	77	56.0	20.1	**0.0270**	0.140	518.0	ug/L	48	Standard
>	**Ga**	71	48.3	15.8				mg/L	90	Standard
	Rb	85	23.3	32.7				ug/L	23	Standard
	Y	89	442158.0	1.4				ug/L	448510	Standard
>	**Rh**	103	8.3	69.3				ug/L	15	Standard
	Mo	98	390.5	34.5	**0.1230**	0.050	40.8	ug/L	11	Standard
	Ag	107	293.7	71.4	**0.0269**	0.031	115.5	ug/L	134	Standard
	Cd	111	93.5	118.9	**0.0403**	0.052	128.0	mg/L	7	Standard
	Cd	114	191.9	127.3	**0.0316**	0.047	147.6	ug/L	16	Standard
>	**In**	115	622178.9	0.7				ug/L	611117	Standard
	Sn	118	798.4	19.1	**0.0306**	0.025	81.3	ug/L	518	Standard
	Sb	123	1327.2	22.5	**0.2633**	0.061	23.3	ug/L	127	Standard
	Ba	135	121.3	95.4	**0.0329**	0.056	169.3	ug/L	45	Standard
	Ce	140	156.7	109.9				ug/L	298	Standard
>	**Tb**	159	979098.8	1.1				ug/L	977942	Standard
	Ho	165	8.3	124.9				ug/L	12	Standard
	Tl	203	319.0	100.9	**0.0316**	0.036	115.5	ug/L	31	Standard
	Tl	205	186.7	86.1	**0.0222**	0.021	94.2	ug/L	20	Standard
	Pb	206	461.3	35.4	**0.0186**	0.029	153.9	ug/L	286	Standard
	Pb	207	392.3	40.1	**0.0194**	0.031	157.4	ug/L	257	Standard
	Pb	208	1505.0	29.2	**0.0164**	0.022	134.2	ug/L	956	Standard
	U	238	157.3	69.9	**0.0181**	0.015	80.5	ug/L	17	Standard
>	**Bi**	209	578116.8	1.4				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:15:29
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	31.7	18.2	**-0.0152**	0.009	57.6	mg/L	32	Standard
K	39	23.3	49.5	**-0.0625**	0.040	64.6	mg/L	30	Standard
Ca	43	25.0	40.0	**-1.6032**	1.411	88.0	mg/L	48	Standard
Fe	54	127.6	20.0	**-0.0223**	0.025	111.2	mg/L	159	Standard
Fe	57	173.3	7.3	**-0.0450**	0.038	85.1	mg/L	195	Standard
> **Sc-1**	45	39924.2	3.2				mg/L	39543	Standard
Cl	35	32496.1	0.6				ug/L	27642	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	1573.4	6.5				ug/L	1647	Standard
P	31	17568.5	2.8				ug/L	15231	Standard
S	34	2373.5	3.0				ug/L	1845	Standard
Sr	88	50.0	26.5				ug/L	70	Standard
C	12	106.7	19.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	12.5	38.4				mg/L	12	Standard
Ho-1	165	8.3	124.9				mg/L	12	Standard
Er	166	16.7	124.9				mg/L	20	Standard
I	127	1251.7	20.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.505	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:15:29
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	101.810
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	103.776
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Sb	123	

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:15:29
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Tuesday, June 28, 2016 15:16:29
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81011.1	2.2				ug/L	81760	Standard
	Be	9	195.0	20.5	**0.1880**	0.037	19.7	ug/L	17	Standard
	Al	27	1493.4	21.7	**-0.0047**	0.002	41.0	ug/L	1410	Standard
	Sc	45	36542.4	1.6				ug/L	39543	Standard
	Ti	47	30.7	9.4	**-0.0369**	0.011	30.3	ug/L	39	Standard
	V	51	3567.1	2.8	**0.3274**	0.017	5.3	ug/L	1415	Standard
	Cr	52	11052.7	0.9	**0.5951**	0.027	4.6	ug/L	6946	Standard
	Cr	53	1095.0	4.8	**0.9563**	0.056	5.9	ug/L	328	Standard
	Mn	55	3466.7	0.6	**0.4079**	0.007	1.6	ug/L	823	Standard
	Co	59	2834.6	2.1	**0.3884**	0.006	1.6	ug/L	252	Standard
	Ni	60	2871.9	0.5	**1.4818**	0.019	1.3	ug/L	192	Standard
	Cu	65	1515.4	2.2	**0.7391**	0.023	3.1	ug/L	141	Standard
	Zn	66	7448.8	1.9	**6.2435**	0.137	2.2	ug/L	491	Standard
>	**Ge**	72	495055.6	0.7				ug/L	515087	Standard
	As	75	363.8	8.2	**0.3769**	0.025	6.6	ug/L	20	Standard
	Se	82	58.7	3.6	**0.4109**	0.015	3.7	ug/L	15	Standard
	Se-1	77	71.3	17.1	**0.2665**	0.171	64.2	ug/L	48	Standard
>	**Ga**	71	38.3	19.9				mg/L	90	Standard
	Rb	85	26.7	28.6				ug/L	23	Standard
	Y	89	415588.9	2.9				ug/L	448510	Standard
>	**Rh**	103	6.7	43.3				ug/L	15	Standard
	Mo	98	101.2	14.2	**0.0199**	0.005	26.4	ug/L	11	Standard
	Ag	107	2581.2	2.0	**0.3840**	0.010	2.6	ug/L	134	Standard
	Cd	111	489.5	5.9	**0.2354**	0.013	5.3	mg/L	7	Standard
	Cd	114	1118.1	7.7	**0.2189**	0.015	6.9	ug/L	16	Standard
>	**In**	115	587241.6	1.2				ug/L	611117	Standard
	Sn	118	508.3	8.2	**-0.0124**	0.008	65.2	ug/L	518	Standard
	Sb	123	2001.1	5.7	**0.4288**	0.031	7.2	ug/L	127	Standard
	Ba	135	1392.4	6.1	**0.6767**	0.039	5.8	ug/L	45	Standard
	Ce	140	30.0	44.1				ug/L	298	Standard
>	**Tb**	159	929212.6	0.4				ug/L	977942	Standard
	Ho	165	13.3	57.3				ug/L	12	Standard
	Tl	203	711.0	4.8	**0.0804**	0.004	5.1	ug/L	31	Standard
	Tl	205	656.7	5.8	**0.0883**	0.005	5.1	ug/L	20	Standard
	Pb	206	1347.1	1.6	**0.1880**	0.004	2.0	ug/L	286	Standard
	Pb	207	1117.4	2.9	**0.1733**	0.005	3.1	ug/L	257	Standard
	Pb	208	4713.3	2.2	**0.1919**	0.007	3.7	ug/L	956	Standard
	U	238	2843.3	0.8	**0.3977**	0.005	1.3	ug/L	17	Standard
>	**Bi**	209	551791.7	0.8				ug/L	557080	Standard

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 15:18:46
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	26.7	39.0	**-0.0189**	0.015	77.3	mg/L	32	Standard
K	39	20.0	25.0	**-0.0684**	0.018	25.7	mg/L	30	Standard
Ca	43	25.0	40.0	**-1.2235**	1.703	139.2	mg/L	48	Standard
Fe	54	112.3	6.9	**-0.0270**	0.008	31.0	mg/L	159	Standard
Fe	57	200.0	28.8	**0.1050**	0.205	195.5	mg/L	195	Standard
Sc-1	45	36542.4	1.6				mg/L	39543	Standard
Cl	35	31755.9	1.1				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1523.4	8.9				ug/L	1647	Standard
P	31	12263.3	0.6				ug/L	15231	Standard
S	34	2283.5	2.6				ug/L	1845	Standard
Sr	88	63.3	25.4				ug/L	70	Standard
C	12	146.7	38.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	15.9	37.5				mg/L	12	Standard
Ho-1	165	13.3	57.3				mg/L	12	Standard
Er	166	16.7	91.7				mg/L	20	Standard
I	127	1103.4	23.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	94.012		
Al	27			
Sc	45			
Ti	47			
V	51	81.855		
Cr	52	74.385		
Cr	53			
Mn	55	81.586		
Co	59	97.096		
Ni	60	92.611		
Cu	65	92.383		
Zn	66	99.896		
Ge	72		96.111	
As	75	94.213		
Se	82	102.730		
Se-1	77			
Ga	71			

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 15:18:46
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	96.001		
Cd	111	98.068		
Cd	114			
> In	115			96.093
Sn	118			
Sb	123	107.193		
Ba	135	90.223		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	100.519		
Tl	205			
Pb	206			
Pb	207			
Pb	208	95.932		
U	238	99.422		
> Bi	209			99.051
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 15:18:46
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078703 WG573333-01
Sample Date/Time: Tuesday, June 28, 2016 15:24:24
Number of Replicates: 3
Autosampler Position: 223
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77973.6	3.5				ug/L	81760	Standard
	Be	9	93.3	32.7	**0.0882**	0.029	33.1	ug/L	17	Standard
	Al	27	522182.4	1.6	**3.5410**	0.149	4.2	ug/L	1410	Standard
	Sc	45	36268.4	3.8				ug/L	39543	Standard
	Ti	47	171.7	14.8	**0.5849**	0.118	20.2	ug/L	39	Standard
	V	51	8345.2	3.6	**1.0965**	0.041	3.8	ug/L	1415	Standard
	Cr	52	9135.4	1.3	**0.3675**	0.032	8.6	ug/L	6946	Standard
	Cr	53	715.0	6.0	**0.5424**	0.045	8.2	ug/L	328	Standard
	Mn	55	230272.3	2.4	**38.4029**	0.677	1.8	ug/L	823	Standard
	Co	59	4116.2	4.4	**0.6044**	0.024	3.9	ug/L	252	Standard
	Ni	60	2251.2	4.1	**1.1925**	0.054	4.6	ug/L	192	Standard
	Cu	65	1784.1	4.6	**0.9285**	0.051	5.5	ug/L	141	Standard
	Zn	66	3508.4	21.7	**2.6509**	0.786	29.6	ug/L	491	Standard
>	**Ge**	72	474556.6	0.9				ug/L	515087	Standard
	As	75	172.3	13.5	**0.2068**	0.024	11.5	ug/L	20	Standard
	Se	82	19.4	8.6	**0.0638**	0.017	27.3	ug/L	15	Standard
	Se-1	77	47.7	3.2	**-0.0320**	0.028	88.4	ug/L	48	Standard
>	**Ga**	71	883.4	11.6				mg/L	90	Standard
	Rb	85	7710.3	5.5				ug/L	23	Standard
	Y	89	442484.5	1.8				ug/L	448510	Standard
>	**Rh**	103	20.0	25.0				ug/L	15	Standard
	Mo	98	50.0	20.9	**0.0005**	0.004	864.7	ug/L	11	Standard
	Ag	107	103.7	35.8	**0.0001**	0.006	5294.7	ug/L	134	Standard
	Cd	111	33.9	43.4	**0.0139**	0.007	52.6	mg/L	7	Standard
	Cd	114	75.2	57.1	**0.0104**	0.009	84.2	ug/L	16	Standard
>	**In**	115	577946.8	0.2				ug/L	611117	Standard
	Sn	118	366.7	11.4	**-0.0364**	0.008	20.8	ug/L	518	Standard
	Sb	123	219.5	27.2	**0.0364**	0.013	37.0	ug/L	127	Standard
	Ba	135	15089.5	2.8	**7.7036**	0.230	3.0	ug/L	45	Standard
	Ce	140	101663.3	0.5				ug/L	298	Standard
>	**Tb**	159	912358.4	0.5				ug/L	977942	Standard
	Ho	165	1823.4	12.2				ug/L	12	Standard
	Tl	203	96.3	39.3	**0.0070**	0.005	65.1	ug/L	31	Standard
	Tl	205	60.0	22.0	**0.0062**	0.002	29.8	ug/L	20	Standard
	Pb	206	5271.9	0.9	**0.9471**	0.013	1.4	ug/L	286	Standard
	Pb	207	4329.0	2.2	**0.8620**	0.023	2.7	ug/L	257	Standard
	Pb	208	18126.6	1.5	**0.9354**	0.019	2.1	ug/L	956	Standard
	U	238	313.3	4.9	**0.0423**	0.002	4.7	ug/L	17	Standard
>	**Bi**	209	537668.4	0.5				ug/L	557080	Standard

Sample ID: L1606078703 WG573333-01
Report Date/Time: Tuesday, June 28, 2016 15:26:41
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	30.0	60.1	**-0.0138**	0.025	183.5	mg/L	32	Standard
K	39	53.3	37.9	**0.0534**	0.067	124.6	mg/L	30	Standard
Ca	43	48.3	6.0	**2.6636**	0.532	20.0	mg/L	48	Standard
Fe	54	632.5	2.4	**0.4873**	0.016	3.2	mg/L	159	Standard
Fe	57	343.3	3.0	**0.6393**	0.010	1.6	mg/L	195	Standard
> **Sc-1**	45	36268.4	3.8				mg/L	39543	Standard
Cl	35	30788.6	1.3				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	1480.1	1.8				ug/L	1647	Standard
P	31	7060.0	1.2				ug/L	15231	Standard
S	34	2285.2	6.0				ug/L	1845	Standard
Sr	88	48.3	21.5				ug/L	70	Standard
C	12	146.7	76.6				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	3101.1	5.7				mg/L	12	Standard
Ho-1	165	1823.4	12.2				mg/L	12	Standard
Er	166	1723.4	3.7				mg/L	20	Standard
I	127	1973.5	19.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		95.369	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.131	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078703 WG573333-01
Report Date/Time: Tuesday, June 28, 2016 15:26:41
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.572
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.515
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078704S WG573333-05
Sample Date/Time: Tuesday, June 28, 2016 15:27:35
Number of Replicates: 3
Autosampler Position: 224
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77117.5	2.8				ug/L	81760	Standard
	Be	9	2866.9	4.2	**3.0501**	0.044	1.4	ug/L	17	Standard
	Al	27	530634.0	2.5	**3.6356**	0.032	0.9	ug/L	1410	Standard
	Sc	45	36769.6	1.1				ug/L	39543	Standard
	Ti	47	373.0	7.6	**1.4053**	0.111	7.9	ug/L	39	Standard
	V	51	27357.6	1.0	**3.9153**	0.028	0.7	ug/L	1415	Standard
	Cr	52	27754.1	2.5	**3.1178**	0.072	2.3	ug/L	6946	Standard
	Cr	53	3205.3	2.9	**3.4822**	0.093	2.7	ug/L	328	Standard
	Mn	55	199454.4	1.6	**32.0700**	0.433	1.4	ug/L	823	Standard
	Co	59	21527.3	1.4	**3.1772**	0.038	1.2	ug/L	252	Standard
	Ni	60	6879.6	1.3	**3.7203**	0.017	0.5	ug/L	192	Standard
	Cu	65	6482.7	2.3	**3.4817**	0.050	1.4	ug/L	141	Standard
	Zn	66	6297.6	0.7	**5.1896**	0.074	1.4	ug/L	491	Standard
>	**Ge**	72	491870.4	0.9				ug/L	515087	Standard
	As	75	2731.5	2.2	**2.5846**	0.063	2.4	ug/L	20	Standard
	Se	82	265.3	2.1	**2.2945**	0.068	3.0	ug/L	15	Standard
	Se-1	77	206.7	1.5	**2.1542**	0.060	2.8	ug/L	48	Standard
>	**Ga**	71	993.4	9.7				mg/L	90	Standard
	Rb	85	7418.5	3.4				ug/L	23	Standard
	Y	89	432036.7	0.9				ug/L	448510	Standard
>	**Rh**	103	6.7	86.6				ug/L	15	Standard
	Mo	98	97.6	12.4	**0.0190**	0.004	23.4	ug/L	11	Standard
	Ag	107	16692.5	1.2	**2.6046**	0.053	2.0	ug/L	134	Standard
	Cd	111	5558.2	2.0	**2.7376**	0.075	2.7	mg/L	7	Standard
	Cd	114	13660.0	0.7	**2.7659**	0.042	1.5	ug/L	16	Standard
>	**In**	115	579914.2	0.8				ug/L	611117	Standard
	Sn	118	355.0	9.2	**-0.0387**	0.005	13.8	ug/L	518	Standard
	Sb	123	2408.5	2.5	**0.5252**	0.017	3.2	ug/L	127	Standard
	Ba	135	18788.0	0.3	**9.5654**	0.079	0.8	ug/L	45	Standard
	Ce	140	97333.3	0.8				ug/L	298	Standard
>	**Tb**	159	916801.1	0.1				ug/L	977942	Standard
	Ho	165	2003.5	4.3				ug/L	12	Standard
	Tl	203	21211.2	1.5	**2.5655**	0.023	0.9	ug/L	31	Standard
	Tl	205	18352.8	1.2	**2.5587**	0.022	0.9	ug/L	20	Standard
	Pb	206	18796.3	1.3	**3.4874**	0.034	1.0	ug/L	286	Standard
	Pb	207	16054.8	0.9	**3.3044**	0.015	0.5	ug/L	257	Standard
	Pb	208	66020.8	0.7	**3.5136**	0.008	0.2	ug/L	956	Standard
	U	238	18577.7	1.3	**2.6444**	0.019	0.7	ug/L	17	Standard
>	**Bi**	209	545605.1	0.6				ug/L	557080	Standard

Sample ID: L1606078704S WG573333-05
Report Date/Time: Tuesday, June 28, 2016 15:29:52
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**23.0608**	19.970	86.6	mg/L	0	Standard
Mg	24	33.3	17.3	**-0.0095**	0.008	83.3	mg/L	32	Standard
K	39	50.0	30.0	**0.0407**	0.057	139.0	mg/L	30	Standard
Ca	43	23.3	86.6	**-1.5565**	3.262	209.6	mg/L	48	Standard
Fe	54	789.3	5.9	**0.6315**	0.054	8.6	mg/L	159	Standard
Fe	57	388.3	19.5	**0.7873**	0.291	37.0	mg/L	195	Standard
> **Sc-1**	45	36769.6	1.1				mg/L	39543	Standard
Cl	35	31616.3	0.3				ug/L	27642	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	1516.7	10.3				ug/L	1647	Standard
P	31	7133.3	2.5				ug/L	15231	Standard
S	34	2288.5	0.7				ug/L	1845	Standard
Sr	88	63.3	12.1				ug/L	70	Standard
C	12	90.0	50.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	3099.8	4.3				mg/L	12	Standard
Ho-1	165	2003.5	4.3				mg/L	12	Standard
Er	166	1750.1	2.1				mg/L	20	Standard
I	127	2031.8	18.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.322	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.493	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078704S WG573333-05
Report Date/Time: Tuesday, June 28, 2016 15:29:52
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.894
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.940
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078705SD WG573333-06
Sample Date/Time: Tuesday, June 28, 2016 15:30:46
Number of Replicates: 3
Autosampler Position: 225
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74501.2	2.9				ug/L	81760	Standard
	Be	9	2816.9	5.3	**3.1033**	0.143	4.6	ug/L	17	Standard
	Al	27	652490.2	1.4	**4.6325**	0.073	1.6	ug/L	1410	Standard
	Sc	45	35373.0	4.0				ug/L	39543	Standard
	Ti	47	288.0	4.5	**1.0503**	0.068	6.5	ug/L	39	Standard
	V	51	27556.1	4.5	**3.9507**	0.115	2.9	ug/L	1415	Standard
	Cr	52	29741.5	2.7	**3.4242**	0.036	1.1	ug/L	6946	Standard
	Cr	53	3408.7	4.0	**3.7308**	0.095	2.5	ug/L	328	Standard
	Mn	55	239390.4	1.7	**38.5872**	0.363	0.9	ug/L	823	Standard
	Co	59	21175.8	2.9	**3.1300**	0.064	2.1	ug/L	252	Standard
	Ni	60	7152.0	2.5	**3.8782**	0.013	0.3	ug/L	192	Standard
	Cu	65	6817.5	3.4	**3.6719**	0.046	1.3	ug/L	141	Standard
	Zn	66	6491.4	2.6	**5.3843**	0.067	1.2	ug/L	491	Standard
>	**Ge**	72	491074.4	2.2				ug/L	515087	Standard
	As	75	2669.0	1.7	**2.5305**	0.018	0.7	ug/L	20	Standard
	Se	82	259.6	0.9	**2.2467**	0.073	3.2	ug/L	15	Standard
	Se-1	77	214.7	9.1	**2.2690**	0.247	10.9	ug/L	48	Standard
>	**Ga**	71	1028.4	8.9				mg/L	90	Standard
	Rb	85	7220.1	2.0				ug/L	23	Standard
	Y	89	434100.4	1.8				ug/L	448510	Standard
>	**Rh**	103	11.7	24.7				ug/L	15	Standard
	Mo	98	220.7	3.6	**0.0685**	0.004	5.3	ug/L	11	Standard
	Ag	107	16241.0	2.2	**2.5784**	0.042	1.6	ug/L	134	Standard
	Cd	111	5563.7	4.0	**2.7880**	0.091	3.3	mg/L	7	Standard
	Cd	114	13468.2	1.0	**2.7750**	0.021	0.8	ug/L	16	Standard
>	**In**	115	569849.5	0.7				ug/L	611117	Standard
	Sn	118	353.3	16.1	**-0.0379**	0.010	26.3	ug/L	518	Standard
	Sb	123	2719.7	5.6	**0.6052**	0.030	5.0	ug/L	127	Standard
	Ba	135	20622.4	3.5	**10.6862**	0.314	2.9	ug/L	45	Standard
	Ce	140	98046.3	2.6				ug/L	298	Standard
>	**Tb**	159	907491.8	2.5				ug/L	977942	Standard
	Ho	165	2030.1	2.1				ug/L	12	Standard
	Tl	203	20792.6	2.1	**2.5412**	0.024	1.0	ug/L	31	Standard
	Tl	205	18564.7	0.9	**2.6157**	0.018	0.7	ug/L	20	Standard
	Pb	206	19134.8	4.2	**3.5891**	0.137	3.8	ug/L	286	Standard
	Pb	207	16318.5	3.9	**3.3949**	0.097	2.9	ug/L	257	Standard
	Pb	208	68373.0	0.9	**3.6800**	0.027	0.7	ug/L	956	Standard
	U	238	18425.5	1.4	**2.6505**	0.039	1.5	ug/L	17	Standard
>	**Bi**	209	539937.0	1.4				ug/L	557080	Standard

Sample ID: L1606078705SD WG573333-06
Report Date/Time: Tuesday, June 28, 2016 15:33:03
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	30.0	44.1	**-0.0120**	0.022	182.1	mg/L	32	Standard
K	39	43.3	29.0	**0.0214**	0.044	203.2	mg/L	30	Standard
Ca	43	31.7	50.8	**-0.0240**	2.524	10500.7	mg/L	48	Standard
Fe	54	837.2	3.3	**0.7115**	0.061	8.6	mg/L	159	Standard
Fe	57	390.0	5.9	**0.8507**	0.142	16.6	mg/L	195	Standard
> **Sc-1**	45	35373.0	4.0				mg/L	39543	Standard
Cl	35	30539.4	1.3				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	1456.7	10.3				ug/L	1647	Standard
P	31	7050.0	3.7				ug/L	15231	Standard
S	34	2141.8	2.3				ug/L	1845	Standard
Sr	88	61.7	12.4				ug/L	70	Standard
C	12	110.0	32.8				mg/L	113	Standard
N	14	6.7	86.6				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	3163.0	6.4				mg/L	12	Standard
Ho-1	165	2030.1	2.1				mg/L	12	Standard
Er	166	1753.4	8.6				mg/L	20	Standard
I	127	1965.1	14.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.122	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.338	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078705SD WG573333-06
Report Date/Time: Tuesday, June 28, 2016 15:33:03
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		93.247
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.923
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078706
Sample Date/Time: Tuesday, June 28, 2016 15:33:58
Number of Replicates: 3
Autosampler Position: 226
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76028.8	3.5				ug/L	81760	Standard
	Be	9	38.3	19.9	**0.0318**	0.009	27.2	ug/L	17	Standard
	Al	27	238262.4	1.5	**1.6489**	0.050	3.0	ug/L	1410	Standard
	Sc	45	37198.9	1.1				ug/L	39543	Standard
	Ti	47	135.7	11.8	**0.3914**	0.065	16.6	ug/L	39	Standard
	V	51	5276.5	1.4	**0.5679**	0.012	2.1	ug/L	1415	Standard
	Cr	52	6498.4	1.5	**-0.1047**	0.015	14.1	ug/L	6946	Standard
	Cr	53	391.7	9.7	**0.1138**	0.045	39.3	ug/L	328	Standard
	Mn	55	119285.7	3.0	**18.6233**	0.543	2.9	ug/L	823	Standard
	Co	59	1493.7	2.3	**0.1855**	0.005	2.9	ug/L	252	Standard
	Ni	60	894.4	1.7	**0.3795**	0.008	2.0	ug/L	192	Standard
	Cu	65	697.3	2.0	**0.2836**	0.007	2.5	ug/L	141	Standard
	Zn	66	2625.6	1.0	**1.6163**	0.028	1.8	ug/L	491	Standard
>	**Ge**	72	504872.7	0.1				ug/L	515087	Standard
	As	75	131.9	5.5	**0.1599**	0.007	4.3	ug/L	20	Standard
	Se	82	16.6	26.9	**0.0277**	0.040	143.8	ug/L	15	Standard
	Se-1	77	42.3	11.2	**-0.1455**	0.065	44.4	ug/L	48	Standard
>	**Ga**	71	331.7	11.3				mg/L	90	Standard
	Rb	85	2800.3	5.6				ug/L	23	Standard
	Y	89	425969.9	0.3				ug/L	448510	Standard
>	**Rh**	103	21.7	96.1				ug/L	15	Standard
	Mo	98	33.2	18.6	**-0.0062**	0.003	40.5	ug/L	11	Standard
	Ag	107	87.7	13.4	**-0.0025**	0.002	77.6	ug/L	134	Standard
	Cd	111	11.3	20.5	**0.0026**	0.001	43.1	mg/L	7	Standard
	Cd	114	30.9	19.2	**0.0013**	0.001	87.3	ug/L	16	Standard
>	**In**	115	583702.0	1.1				ug/L	611117	Standard
	Sn	118	278.3	22.7	**-0.0527**	0.011	20.3	ug/L	518	Standard
	Sb	123	103.7	16.2	**0.0102**	0.004	35.3	ug/L	127	Standard
	Ba	135	9187.1	1.4	**4.6338**	0.016	0.3	ug/L	45	Standard
	Ce	140	45456.4	1.3				ug/L	298	Standard
>	**Tb**	159	930779.6	1.7				ug/L	977942	Standard
	Ho	165	850.0	2.4				ug/L	12	Standard
	Tl	203	23.3	23.6	**-0.0020**	0.001	31.7	ug/L	31	Standard
	Tl	205	35.0	37.8	**0.0025**	0.002	71.7	ug/L	20	Standard
	Pb	206	2536.5	1.2	**0.4062**	0.009	2.2	ug/L	286	Standard
	Pb	207	2050.5	1.7	**0.3630**	0.010	2.7	ug/L	257	Standard
	Pb	208	8421.9	1.7	**0.3865**	0.010	2.6	ug/L	956	Standard
	U	238	103.3	13.1	**0.0115**	0.002	17.1	ug/L	17	Standard
>	**Bi**	209	556617.5	0.8				ug/L	557080	Standard

Sample ID: L1606078706
Report Date/Time: Tuesday, June 28, 2016 15:36:15
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**11.3116**	19.584	173.1	mg/L	0	Standard
Mg	24	31.7	9.1	**-0.0123**	0.005	37.6	mg/L	32	Standard
K	39	50.0	10.0	**0.0383**	0.020	51.6	mg/L	30	Standard
Ca	43	18.3	103.3	**-2.3989**	3.033	126.4	mg/L	48	Standard
Fe	54	317.8	14.7	**0.1684**	0.043	25.3	mg/L	159	Standard
Fe	57	280.0	19.3	**0.3807**	0.196	51.6	mg/L	195	Standard
> **Sc-1**	45	37198.9	1.1				mg/L	39543	Standard
Cl	35	30150.0	0.9				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	1330.1	6.0				ug/L	1647	Standard
P	31	7115.0	3.6				ug/L	15231	Standard
S	34	2203.5	6.4				ug/L	1845	Standard
Sr	88	73.3	23.9				ug/L	70	Standard
C	12	136.7	33.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	1461.3	2.6				mg/L	12	Standard
Ho-1	165	850.0	2.4				mg/L	12	Standard
Er	166	740.0	2.3				mg/L	20	Standard
I	127	1331.7	15.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		92.990	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.017	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078706
Report Date/Time: Tuesday, June 28, 2016 15:36:15
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.514
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.917
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078708
Sample Date/Time: Tuesday, June 28, 2016 15:37:09
Number of Replicates: 3
Autosampler Position: 227
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75812.6	2.8				ug/L	81760	Standard
	Be	9	48.3	21.5	**0.0428**	0.011	26.6	ug/L	17	Standard
	Al	27	1071474.1	1.0	**7.4874**	0.284	3.8	ug/L	1410	Standard
	Sc	45	37513.1	2.1				ug/L	39543	Standard
	Ti	47	158.0	12.8	**0.5014**	0.093	18.5	ug/L	39	Standard
	V	51	6689.1	0.4	**0.8020**	0.010	1.2	ug/L	1415	Standard
	Cr	52	7515.2	0.4	**0.0743**	0.008	10.8	ug/L	6946	Standard
	Cr	53	588.3	8.1	**0.3608**	0.056	15.4	ug/L	328	Standard
	Mn	55	205213.7	2.8	**33.0324**	1.280	3.9	ug/L	823	Standard
	Co	59	2212.5	3.2	**0.2987**	0.013	4.3	ug/L	252	Standard
	Ni	60	1404.4	3.4	**0.6767**	0.035	5.1	ug/L	192	Standard
	Cu	65	1102.0	3.5	**0.5169**	0.018	3.4	ug/L	141	Standard
	Zn	66	2607.9	0.9	**1.6660**	0.041	2.5	ug/L	491	Standard
>	**Ge**	72	491532.3	1.1				ug/L	515087	Standard
	As	75	127.5	11.1	**0.1592**	0.014	8.7	ug/L	20	Standard
	Se	82	17.3	29.3	**0.0385**	0.047	123.2	ug/L	15	Standard
	Se-1	77	43.3	11.6	**-0.1157**	0.075	64.9	ug/L	48	Standard
>	**Ga**	71	450.0	13.5				mg/L	90	Standard
	Rb	85	3788.8	1.3				ug/L	23	Standard
	Y	89	433725.3	2.1				ug/L	448510	Standard
>	**Rh**	103	18.3	31.5				ug/L	15	Standard
	Mo	98	38.5	4.2	**-0.0045**	0.000	10.9	ug/L	11	Standard
	Ag	107	93.3	12.0	**-0.0020**	0.002	78.2	ug/L	134	Standard
	Cd	111	23.9	23.2	**0.0085**	0.003	31.3	mg/L	7	Standard
	Cd	114	45.7	16.0	**0.0041**	0.001	34.3	ug/L	16	Standard
>	**In**	115	598611.3	0.9				ug/L	611117	Standard
	Sn	118	346.7	46.9	**-0.0421**	0.028	66.0	ug/L	518	Standard
	Sb	123	86.1	26.8	**0.0058**	0.005	89.2	ug/L	127	Standard
	Ba	135	10124.4	4.3	**4.9826**	0.248	5.0	ug/L	45	Standard
	Ce	140	56269.4	3.6				ug/L	298	Standard
>	**Tb**	159	935742.2	1.5				ug/L	977942	Standard
	Ho	165	1033.4	8.1				ug/L	12	Standard
	Tl	203	29.7	5.1	**-0.0013**	0.000	18.3	ug/L	31	Standard
	Tl	205	31.7	18.2	**0.0020**	0.001	35.8	ug/L	20	Standard
	Pb	206	3055.6	2.0	**0.5003**	0.022	4.4	ug/L	286	Standard
	Pb	207	2503.2	3.8	**0.4542**	0.030	6.6	ug/L	257	Standard
	Pb	208	10181.4	1.0	**0.4779**	0.016	3.3	ug/L	956	Standard
	U	238	166.3	6.9	**0.0202**	0.002	9.9	ug/L	17	Standard
>	**Bi**	209	558847.5	2.0				ug/L	557080	Standard

Sample ID: L1606078708
Report Date/Time: Tuesday, June 28, 2016 15:39:26
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	26.7	28.6	**-0.0198**	0.010	52.5	mg/L	32	Standard
K	39	46.7	22.3	**0.0253**	0.040	158.3	mg/L	30	Standard
Ca	43	46.7	43.3	**2.1703**	3.421	157.6	mg/L	48	Standard
Fe	54	430.9	12.9	**0.2746**	0.059	21.6	mg/L	159	Standard
Fe	57	336.7	7.3	**0.5728**	0.068	12.0	mg/L	195	Standard
Sc-1	45	37513.1	2.1				mg/L	39543	Standard
Cl	35	31262.2	0.8				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	1390.1	8.3				ug/L	1647	Standard
P	31	7195.0	0.6				ug/L	15231	Standard
S	34	2350.2	4.3				ug/L	1845	Standard
Sr	88	86.7	13.3				ug/L	70	Standard
C	12	130.0	15.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	1824.8	9.6				mg/L	12	Standard
Ho-1	165	1033.4	8.1				mg/L	12	Standard
Er	166	946.7	6.2				mg/L	20	Standard
I	127	1363.4	17.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		92.726	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.427	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078708
Report Date/Time: Tuesday, June 28, 2016 15:39:26
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	97.954	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	100.317	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078713
Sample Date/Time: Tuesday, June 28, 2016 15:40:21
Number of Replicates: 3
Autosampler Position: 228
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75663.5	2.7				ug/L	81760	Standard
	Be	9	81.7	30.2	**0.0793**	0.028	34.9	ug/L	17	Standard
	Al	27	416574.6	1.3	**2.9066**	0.048	1.7	ug/L	1410	Standard
	Sc	45	36378.7	3.4				ug/L	39543	Standard
	Ti	47	126.3	10.5	**0.3773**	0.057	15.2	ug/L	39	Standard
	V	51	8315.3	0.5	**1.0708**	0.007	0.7	ug/L	1415	Standard
	Cr	52	8765.5	1.6	**0.2875**	0.022	7.8	ug/L	6946	Standard
	Cr	53	728.4	7.9	**0.5445**	0.070	12.8	ug/L	328	Standard
	Mn	55	93455.4	1.8	**15.2441**	0.287	1.9	ug/L	823	Standard
	Co	59	3649.1	3.3	**0.5232**	0.019	3.5	ug/L	252	Standard
	Ni	60	1651.1	5.4	**0.8312**	0.052	6.2	ug/L	192	Standard
	Cu	65	1200.0	2.8	**0.5835**	0.019	3.3	ug/L	141	Standard
	Zn	66	2593.9	1.4	**1.6993**	0.037	2.2	ug/L	491	Standard
>	**Ge**	72	482399.4	0.1				ug/L	515087	Standard
	As	75	167.0	21.6	**0.1989**	0.034	17.3	ug/L	20	Standard
	Se	82	15.9	30.3	**0.0282**	0.045	158.8	ug/L	15	Standard
	Se-1	77	52.3	6.7	**0.0228**	0.050	216.8	ug/L	48	Standard
>	**Ga**	71	573.3	13.4				mg/L	90	Standard
	Rb	85	5119.2	1.4				ug/L	23	Standard
	Y	89	439888.1	1.2				ug/L	448510	Standard
>	**Rh**	103	13.3	78.1				ug/L	15	Standard
	Mo	98	43.4	10.8	**-0.0024**	0.002	79.1	ug/L	11	Standard
	Ag	107	106.7	10.0	**0.0002**	0.002	826.8	ug/L	134	Standard
	Cd	111	22.9	46.1	**0.0082**	0.005	63.6	mg/L	7	Standard
	Cd	114	51.2	44.0	**0.0052**	0.004	84.1	ug/L	16	Standard
>	**In**	115	591171.7	0.9				ug/L	611117	Standard
	Sn	118	328.3	8.7	**-0.0445**	0.005	12.1	ug/L	518	Standard
	Sb	123	88.5	32.5	**0.0066**	0.006	98.7	ug/L	127	Standard
	Ba	135	15979.8	0.8	**7.9763**	0.029	0.4	ug/L	45	Standard
	Ce	140	70630.9	0.7				ug/L	298	Standard
>	**Tb**	159	919155.3	2.1				ug/L	977942	Standard
	Ho	165	1308.4	8.9				ug/L	12	Standard
	Tl	203	43.3	33.8	**0.0005**	0.002	355.9	ug/L	31	Standard
	Tl	205	45.0	58.8	**0.0041**	0.004	93.0	ug/L	20	Standard
	Pb	206	4127.6	1.0	**0.7213**	0.011	1.5	ug/L	286	Standard
	Pb	207	3419.4	2.4	**0.6628**	0.016	2.5	ug/L	257	Standard
	Pb	208	14423.4	1.0	**0.7255**	0.013	1.8	ug/L	956	Standard
	U	238	157.3	5.4	**0.0196**	0.001	6.2	ug/L	17	Standard
>	**Bi**	209	542054.4	0.7				ug/L	557080	Standard

Sample ID: L1606078713
Report Date/Time: Tuesday, June 28, 2016 15:42:38
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	38.3	7.5	**-0.0016**	0.005	296.7	mg/L	32	Standard
K	39	31.7	55.5	**-0.0245**	0.065	264.7	mg/L	30	Standard
Ca	43	18.3	56.8	**-2.3425**	1.620	69.1	mg/L	48	Standard
Fe	54	638.8	9.9	**0.4927**	0.076	15.4	mg/L	159	Standard
Fe	57	368.3	10.9	**0.7248**	0.101	14.0	mg/L	195	Standard
> **Sc-1**	45	36378.7	3.4				mg/L	39543	Standard
Cl	35	31640.3	1.2				ug/L	27642	Standard
Kr	83	2.7	57.3				ug/L	2	Standard
Br	81	1476.7	9.6				ug/L	1647	Standard
P	31	7046.6	3.0				ug/L	15231	Standard
S	34	2278.5	11.1				ug/L	1845	Standard
Sr	88	80.0	28.6				ug/L	70	Standard
C	12	116.7	26.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	2168.9	5.7				mg/L	12	Standard
Ho-1	165	1308.4	8.9				mg/L	12	Standard
Er	166	1140.0	7.8				mg/L	20	Standard
I	127	1385.1	16.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		92.543	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.654	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078713
Report Date/Time: Tuesday, June 28, 2016 15:42:38
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.736
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.303
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078715
Sample Date/Time: Tuesday, June 28, 2016 15:43:32
Number of Replicates: 3
Autosampler Position: 229
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72672.4	3.8				ug/L	81760	Standard
	Be	9	136.7	16.5	**0.1453**	0.027	18.6	ug/L	17	Standard
	Al	27	1143496.8	3.3	**8.3341**	0.197	2.4	ug/L	1410	Standard
	Sc	45	36597.5	3.0				ug/L	39543	Standard
	Ti	47	230.0	1.3	**0.8041**	0.017	2.1	ug/L	39	Standard
	V	51	11413.9	1.0	**1.5144**	0.038	2.5	ug/L	1415	Standard
	Cr	52	15036.8	0.3	**1.2067**	0.029	2.4	ug/L	6946	Standard
	Cr	53	1373.4	5.7	**1.2975**	0.078	6.0	ug/L	328	Standard
	Mn	55	293360.4	0.3	**47.2738**	0.683	1.4	ug/L	823	Standard
	Co	59	6852.2	0.9	**0.9907**	0.018	1.8	ug/L	252	Standard
	Ni	60	4182.9	0.4	**2.2225**	0.021	0.9	ug/L	192	Standard
	Cu	65	2749.2	1.6	**1.4253**	0.010	0.7	ug/L	141	Standard
	Zn	66	5934.8	0.5	**4.8465**	0.046	1.0	ug/L	491	Standard
>	**Ge**	72	491565.9	1.3				ug/L	515087	Standard
	As	75	281.2	9.9	**0.3023**	0.024	7.9	ug/L	20	Standard
	Se	82	16.6	41.5	**0.0313**	0.062	197.6	ug/L	15	Standard
	Se-1	77	52.7	11.4	**0.0136**	0.082	602.5	ug/L	48	Standard
>	**Ga**	71	1726.8	3.8				mg/L	90	Standard
	Rb	85	17453.4	2.4				ug/L	23	Standard
	Y	89	433246.4	2.7				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	40.4	18.8	**-0.0027**	0.003	116.5	ug/L	11	Standard
	Ag	107	118.3	7.6	**0.0030**	0.002	54.7	ug/L	134	Standard
	Cd	111	36.9	12.4	**0.0159**	0.002	13.6	mg/L	7	Standard
	Cd	114	100.1	28.9	**0.0161**	0.006	36.4	ug/L	16	Standard
>	**In**	115	559969.6	1.0				ug/L	611117	Standard
	Sn	118	245.0	14.3	**-0.0567**	0.006	11.1	ug/L	518	Standard
	Sb	123	73.5	14.6	**0.0042**	0.003	63.0	ug/L	127	Standard
	Ba	135	26697.5	0.7	**14.0890**	0.209	1.5	ug/L	45	Standard
	Ce	140	126090.9	1.2				ug/L	298	Standard
>	**Tb**	159	888471.0	0.2				ug/L	977942	Standard
	Ho	165	2013.5	3.3				ug/L	12	Standard
	Tl	203	105.3	16.7	**0.0083**	0.002	26.4	ug/L	31	Standard
	Tl	205	98.3	26.1	**0.0118**	0.004	31.7	ug/L	20	Standard
	Pb	206	7239.7	1.0	**1.3394**	0.022	1.6	ug/L	286	Standard
	Pb	207	5907.2	0.7	**1.2110**	0.009	0.7	ug/L	257	Standard
	Pb	208	24779.9	1.0	**1.3155**	0.008	0.6	ug/L	956	Standard
	U	238	285.0	3.3	**0.0387**	0.001	3.3	ug/L	17	Standard
>	**Bi**	209	531928.0	0.6				ug/L	557080	Standard

Sample ID: L1606078715
Report Date/Time: Tuesday, June 28, 2016 15:45:49
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**11.1992**	19.389	173.1	mg/L	0	Standard
Mg	24	35.0	0.0	**-0.0068**	0.002	22.3	mg/L	32	Standard
K	39	70.0	24.7	**0.1145**	0.065	56.7	mg/L	30	Standard
Ca	43	41.7	48.5	**1.5107**	3.417	226.2	mg/L	48	Standard
Fe	54	1498.6	2.3	**1.3298**	0.075	5.6	mg/L	159	Standard
Fe	57	515.0	3.5	**1.2558**	0.109	8.7	mg/L	195	Standard
> **Sc-1**	45	36597.5	3.0				mg/L	39543	Standard
Cl	35	30785.2	1.5				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	1506.7	9.5				ug/L	1647	Standard
P	31	6974.9	2.8				ug/L	15231	Standard
S	34	2301.8	4.8				ug/L	1845	Standard
Sr	88	70.0	28.6				ug/L	70	Standard
C	12	113.3	31.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	3365.6	4.6				mg/L	12	Standard
Ho-1	165	2013.5	3.3				mg/L	12	Standard
Er	166	1770.1	9.2				mg/L	20	Standard
I	127	1690.1	9.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		88.885	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.434	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078715
Report Date/Time: Tuesday, June 28, 2016 15:45:49
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		91.631
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.485
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078716
Sample Date/Time: Tuesday, June 28, 2016 15:46:44
Number of Replicates: 3
Autosampler Position: 230
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72657.2	3.0				ug/L	81760	Standard
	Be	9	110.0	13.6	**0.1151**	0.019	16.6	ug/L	17	Standard
	Al	27	964168.3	0.2	**7.0283**	0.202	2.9	ug/L	1410	Standard
	Sc	45	36550.7	1.7				ug/L	39543	Standard
	Ti	47	122.3	2.9	**0.3511**	0.018	5.2	ug/L	39	Standard
	V	51	5582.1	3.3	**0.6361**	0.029	4.5	ug/L	1415	Standard
	Cr	52	12113.5	1.9	**0.7686**	0.019	2.4	ug/L	6946	Standard
	Cr	53	1075.0	5.2	**0.9442**	0.085	9.0	ug/L	328	Standard
	Mn	55	296391.3	0.6	**47.8289**	1.096	2.3	ug/L	823	Standard
	Co	59	5108.9	2.5	**0.7319**	0.032	4.3	ug/L	252	Standard
	Ni	60	3102.7	1.4	**1.6241**	0.056	3.4	ug/L	192	Standard
	Cu	65	3067.6	0.7	**1.6034**	0.043	2.7	ug/L	141	Standard
	Zn	66	5597.7	1.7	**4.5323**	0.178	3.9	ug/L	491	Standard
>	**Ge**	72	490975.8	1.9				ug/L	515087	Standard
	As	75	34.9	42.7	**0.0731**	0.015	20.1	ug/L	20	Standard
	Se	82	17.1	13.6	**0.0368**	0.020	54.2	ug/L	15	Standard
	Se-1	77	53.7	6.0	**0.0281**	0.032	114.8	ug/L	48	Standard
>	**Ga**	71	1235.1	6.5				mg/L	90	Standard
	Rb	85	7633.6	0.9				ug/L	23	Standard
	Y	89	431776.8	2.6				ug/L	448510	Standard
>	**Rh**	103	15.0	66.7				ug/L	15	Standard
	Mo	98	19.3	30.7	**-0.0115**	0.002	20.4	ug/L	11	Standard
	Ag	107	114.7	7.5	**0.0018**	0.001	48.3	ug/L	134	Standard
	Cd	111	65.3	10.2	**0.0294**	0.003	8.8	mg/L	7	Standard
	Cd	114	154.6	13.1	**0.0266**	0.003	12.9	ug/L	16	Standard
>	**In**	115	577281.8	2.6				ug/L	611117	Standard
	Sn	118	261.7	10.5	**-0.0551**	0.004	6.8	ug/L	518	Standard
	Sb	123	70.5	12.0	**0.0029**	0.001	50.3	ug/L	127	Standard
	Ba	135	31515.4	1.1	**16.1422**	0.456	2.8	ug/L	45	Standard
	Ce	140	108483.7	1.5				ug/L	298	Standard
>	**Tb**	159	919053.3	0.1				ug/L	977942	Standard
	Ho	165	2146.8	3.5				ug/L	12	Standard
	Tl	203	55.7	15.3	**0.0020**	0.001	51.1	ug/L	31	Standard
	Tl	205	56.7	10.2	**0.0057**	0.001	13.8	ug/L	20	Standard
	Pb	206	8383.7	0.8	**1.5358**	0.026	1.7	ug/L	286	Standard
	Pb	207	6950.9	2.1	**1.4122**	0.046	3.3	ug/L	257	Standard
	Pb	208	28623.6	0.6	**1.5044**	0.028	1.9	ug/L	956	Standard
	U	238	278.7	6.7	**0.0371**	0.002	6.0	ug/L	17	Standard
>	**Bi**	209	540360.9	1.3				ug/L	557080	Standard

Sample ID: L1606078716
Report Date/Time: Tuesday, June 28, 2016 15:49:01
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**23.0232**	19.940	86.6	mg/L	0	Standard
Mg	24	31.7	36.5	**-0.0117**	0.016	134.2	mg/L	32	Standard
K	39	40.0	57.3	**0.0042**	0.082	1950.6	mg/L	30	Standard
Ca	43	45.0	19.2	**2.0338**	1.285	63.2	mg/L	48	Standard
Fe	54	558.7	3.5	**0.4103**	0.028	6.8	mg/L	159	Standard
Fe	57	366.7	19.3	**0.7123**	0.240	33.7	mg/L	195	Standard
> **Sc-1**	45	36550.7	1.7				mg/L	39543	Standard
Cl	35	30424.5	0.4				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	1540.1	8.3				ug/L	1647	Standard
P	31	6878.2	1.3				ug/L	15231	Standard
S	34	2278.5	4.6				ug/L	1845	Standard
Sr	88	71.7	29.0				ug/L	70	Standard
C	12	150.0	33.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	0	Standard
Dy	164	3587.7	5.2				mg/L	12	Standard
Ho-1	165	2146.8	3.5				mg/L	12	Standard
Er	166	1936.8	10.5				mg/L	20	Standard
I	127	1943.5	8.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		88.866	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.319	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078716
Report Date/Time: Tuesday, June 28, 2016 15:49:01
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.463
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.999
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078717
Sample Date/Time: Tuesday, June 28, 2016 15:49:55
Number of Replicates: 3
Autosampler Position: 231
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72923.4	1.2				ug/L	81760	Standard
	Be	9	50.0	43.6	**0.0466**	0.024	52.0	ug/L	17	Standard
	Al	27	245816.4	0.3	**1.7738**	0.027	1.5	ug/L	1410	Standard
	Sc	45	37411.1	1.3				ug/L	39543	Standard
	Ti	47	150.7	7.6	**0.4619**	0.045	9.8	ug/L	39	Standard
	V	51	5992.6	2.2	**0.6857**	0.015	2.1	ug/L	1415	Standard
	Cr	52	7687.3	1.9	**0.0860**	0.027	31.9	ug/L	6946	Standard
	Cr	53	518.3	10.1	**0.2697**	0.061	22.7	ug/L	328	Standard
	Mn	55	119301.5	0.3	**18.8992**	0.085	0.4	ug/L	823	Standard
	Co	59	2188.2	2.9	**0.2910**	0.008	2.8	ug/L	252	Standard
	Ni	60	1191.7	0.5	**0.5500**	0.007	1.2	ug/L	192	Standard
	Cu	65	843.0	2.1	**0.3684**	0.010	2.8	ug/L	141	Standard
	Zn	66	2526.9	1.6	**1.5588**	0.051	3.2	ug/L	491	Standard
>	**Ge**	72	497647.3	0.6				ug/L	515087	Standard
	As	75	96.2	13.8	**0.1289**	0.013	9.7	ug/L	20	Standard
	Se	82	10.6	36.2	**-0.0241**	0.035	144.8	ug/L	15	Standard
	Se-1	77	46.0	14.3	**-0.0871**	0.088	100.8	ug/L	48	Standard
>	**Ga**	71	388.3	9.7				mg/L	90	Standard
	Rb	85	3552.1	4.5				ug/L	23	Standard
	Y	89	408184.3	3.4				ug/L	448510	Standard
>	**Rh**	103	16.7	17.3				ug/L	15	Standard
	Mo	98	62.0	6.5	**0.0053**	0.001	22.5	ug/L	11	Standard
	Ag	107	95.7	11.7	**-0.0010**	0.002	195.1	ug/L	134	Standard
	Cd	111	10.9	9.2	**0.0025**	0.001	23.6	mg/L	7	Standard
	Cd	114	23.2	24.4	**-0.0002**	0.001	811.4	ug/L	16	Standard
>	**In**	115	574185.6	1.8				ug/L	611117	Standard
	Sn	118	255.0	18.7	**-0.0561**	0.008	14.6	ug/L	518	Standard
	Sb	123	68.2	5.6	**0.0025**	0.001	35.8	ug/L	127	Standard
	Ba	135	10138.1	0.2	**5.2024**	0.085	1.6	ug/L	45	Standard
	Ce	140	49457.3	0.9				ug/L	298	Standard
>	**Tb**	159	906877.5	2.2				ug/L	977942	Standard
	Ho	165	853.4	7.8				ug/L	12	Standard
	Tl	203	26.0	16.8	**-0.0016**	0.001	32.4	ug/L	31	Standard
	Tl	205	21.7	26.6	**0.0008**	0.001	106.7	ug/L	20	Standard
	Pb	206	2799.3	1.6	**0.4682**	0.011	2.3	ug/L	286	Standard
	Pb	207	2286.8	2.0	**0.4236**	0.011	2.7	ug/L	257	Standard
	Pb	208	9380.8	1.4	**0.4502**	0.009	2.0	ug/L	956	Standard
	U	238	106.7	2.9	**0.0123**	0.000	3.6	ug/L	17	Standard
>	**Bi**	209	542636.2	0.5				ug/L	557080	Standard

Sample ID: L1606078717
Report Date/Time: Tuesday, June 28, 2016 15:52:12
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**22.4620**	19.449	86.6	mg/L	0	Standard
Mg	24	26.7	21.7	**-0.0196**	0.008	42.2	mg/L	32	Standard
K	39	36.7	7.9	**-0.0105**	0.010	94.8	mg/L	30	Standard
Ca	43	36.7	56.8	**0.5265**	3.309	628.5	mg/L	48	Standard
Fe	54	389.3	2.1	**0.2354**	0.013	5.4	mg/L	159	Standard
Fe	57	280.0	3.1	**0.3744**	0.024	6.3	mg/L	195	Standard
> **Sc-1**	45	37411.1	1.3				mg/L	39543	Standard
Cl	35	30529.4	0.4				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	1530.1	5.4				ug/L	1647	Standard
P	31	7263.4	2.6				ug/L	15231	Standard
S	34	2140.2	5.9				ug/L	1845	Standard
Sr	88	53.3	23.6				ug/L	70	Standard
C	12	100.0	75.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	1448.1	5.9				mg/L	12	Standard
Ho-1	165	853.4	7.8				mg/L	12	Standard
Er	166	736.7	15.9				mg/L	20	Standard
I	127	6059.5	4.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		89.192	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.614	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078717
Report Date/Time: Tuesday, June 28, 2016 15:52:12
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		93.957
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.407
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 15:53:08
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	80037.6	2.2				ug/L	81760	Standard
	Be	9	50415.4	0.4	**51.8632**	1.062	2.0	ug/L	17	Standard
	Al	27	7611642.9	1.2	**50.4371**	0.677	1.3	ug/L	1410	Standard
	Sc	45	41097.3	0.2				ug/L	39543	Standard
	Ti	47	25369.9	1.7	**101.8781**	2.068	2.0	ug/L	39	Standard
	V	51	351293.9	2.1	**50.3394**	0.543	1.1	ug/L	1415	Standard
	Cr	52	361193.1	2.4	**50.8341**	0.636	1.3	ug/L	6946	Standard
	Cr	53	45189.0	2.1	**51.1421**	0.625	1.2	ug/L	328	Standard
	Mn	55	322540.4	1.4	**49.6098**	0.122	0.2	ug/L	823	Standard
	Co	59	356756.8	0.9	**50.7536**	0.285	0.6	ug/L	252	Standard
	Ni	60	97064.3	2.1	**51.4369**	0.357	0.7	ug/L	192	Standard
	Cu	65	97360.2	2.2	**51.1525**	0.591	1.2	ug/L	141	Standard
	Zn	66	56918.1	2.1	**51.1056**	0.446	0.9	ug/L	491	Standard
>	Ge	72	515026.5	1.4				ug/L	515087	Standard
	As	75	56903.6	1.1	**50.6598**	0.179	0.4	ug/L	20	Standard
	Se	82	5856.2	1.8	**50.7515**	0.222	0.4	ug/L	15	Standard
	Se-1	77	3927.8	0.7	**51.4293**	0.511	1.0	ug/L	48	Standard
>	Ga	71	66.7	30.3				mg/L	90	Standard
	Rb	85	760.0	4.6				ug/L	23	Standard
	Y	89	447852.9	1.0				ug/L	448510	Standard
>	Rh	103	25.0	34.6				ug/L	15	Standard
	Mo	98	271272.3	0.9	**98.6404**	1.710	1.7	ug/L	11	Standard
	Ag	107	339950.6	2.5	**49.7752**	1.742	3.5	ug/L	134	Standard
	Cd	111	107240.7	1.0	**49.3177**	0.985	2.0	mg/L	7	Standard
	Cd	114	262943.6	0.6	**49.7461**	0.816	1.6	ug/L	16	Standard
>	In	115	621725.8	1.1				ug/L	611117	Standard
	Sn	118	299388.1	2.0	**49.6643**	0.590	1.2	ug/L	518	Standard
	Sb	123	239448.5	1.8	**49.8807**	1.390	2.8	ug/L	127	Standard
	Ba	135	102748.9	0.7	**48.8974**	0.441	0.9	ug/L	45	Standard
	Ce	140	90.0	5.6				ug/L	298	Standard
>	Tb	159	985013.6	1.3				ug/L	977942	Standard
	Ho	165	16.7	45.8				ug/L	12	Standard
	Tl	203	430595.1	1.2	**49.9408**	0.227	0.5	ug/L	31	Standard
	Tl	205	380549.9	0.8	**50.8302**	0.446	0.9	ug/L	20	Standard
	Pb	206	272066.9	1.5	**49.1343**	0.429	0.9	ug/L	286	Standard
	Pb	207	246215.9	1.8	**49.3013**	0.781	1.6	ug/L	257	Standard
	Pb	208	985827.4	0.9	**51.0241**	0.323	0.6	ug/L	956	Standard
	U	238	363383.0	0.6	**49.5660**	0.508	1.0	ug/L	17	Standard
>	Bi	209	570034.9	1.5				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:55:25
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.2894**	17.813	173.1	mg/L	0	Standard
Mg	24	3815.5	2.6	**4.8299**	0.134	2.8	mg/L	32	Standard
K	39	1281.7	1.9	**4.0254**	0.078	1.9	mg/L	30	Standard
Ca	43	63.3	19.9	**3.9099**	1.840	47.1	mg/L	48	Standard
Fe	54	5945.2	1.8	**5.0400**	0.102	2.0	mg/L	159	Standard
Fe	57	1733.4	4.2	**5.0061**	0.238	4.8	mg/L	195	Standard
> **Sc-1**	45	41097.3	0.2				mg/L	39543	Standard
Cl	35	35020.4	1.1				ug/L	27642	Standard
Kr	83	3.0	57.7				ug/L	2	Standard
Br	81	1666.8	6.4				ug/L	1647	Standard
P	31	18748.3	2.5				ug/L	15231	Standard
S	34	2583.6	5.3				ug/L	1845	Standard
Sr	88	53.3	28.6				ug/L	70	Standard
C	12	110.0	24.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.4	101.7				mg/L	12	Standard
Ho-1	165	16.7	45.8				mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	1388.4	22.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	103.726		
	Al	27	100.874		
	Sc	45			
	Ti	47	101.878		
	V	51	100.679		
	Cr	52	101.668		
	Cr	53			
	Mn	55	99.220		
	Co	59	101.507		
	Ni	60	102.874		
	Cu	65	102.305		
	Zn	66	102.211		
>	Ge	72		99.988	
	As	75	101.320		
	Se	82	101.503		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:55:25
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	98.640		
Ag	107	99.550		
Cd	111	98.635		
Cd	114			
In	115			101.736
Sn	118	99.329		
Sb	123	99.761		
Ba	135	97.795		
Ce	140			
Tb	159			
Ho	165			
Tl	203	99.882		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.048		
U	238	99.132		
Bi	209			102.325
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:55:25
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 15:56:20
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79843.2	1.6				ug/L	81760	Standard
	Be	9	23.3	75.3	**0.0144**	0.018	125.7	ug/L	17	Standard
	Al	27	2505.2	18.4	**0.0022**	0.003	136.7	ug/L	1410	Standard
	Sc	45	39966.0	1.7				ug/L	39543	Standard
	Ti	47	35.3	11.4	**-0.0209**	0.017	83.6	ug/L	39	Standard
	V	51	1133.0	9.8	**-0.0412**	0.016	39.7	ug/L	1415	Standard
	Cr	52	6206.9	2.0	**-0.1524**	0.017	11.1	ug/L	6946	Standard
	Cr	53	315.0	2.7	**0.0227**	0.013	56.9	ug/L	328	Standard
	Mn	55	686.0	7.5	**-0.0412**	0.007	17.6	ug/L	823	Standard
	Co	59	254.3	2.2	**0.0054**	0.001	12.6	ug/L	252	Standard
	Ni	60	171.3	1.2	**-0.0127**	0.001	8.3	ug/L	192	Standard
	Cu	65	123.3	12.6	**-0.0250**	0.008	31.0	ug/L	141	Standard
	Zn	66	449.3	3.1	**-0.4117**	0.013	3.2	ug/L	491	Standard
>	**Ge**	72	507665.8	0.8				ug/L	515087	Standard
	As	75	-25.2	101.2	**0.0174**	0.023	132.3	ug/L	20	Standard
	Se	82	17.2	13.7	**0.0322**	0.020	60.6	ug/L	15	Standard
	Se-1	77	51.3	10.7	**-0.0277**	0.072	260.3	ug/L	48	Standard
>	**Ga**	71	48.3	15.8				mg/L	90	Standard
	Rb	85	30.0	44.1				ug/L	23	Standard
	Y	89	444982.9	1.9				ug/L	448510	Standard
>	**Rh**	103	16.7	34.6				ug/L	15	Standard
	Mo	98	252.7	24.3	**0.0725**	0.022	30.7	ug/L	11	Standard
	Ag	107	129.7	3.6	**0.0027**	0.001	26.2	ug/L	134	Standard
	Cd	111	17.2	9.0	**0.0050**	0.001	14.0	mg/L	7	Standard
	Cd	114	47.7	10.5	**0.0041**	0.001	23.1	ug/L	16	Standard
>	**In**	115	623883.8	0.1				ug/L	611117	Standard
	Sn	118	770.0	14.1	**0.0256**	0.018	70.0	ug/L	518	Standard
	Sb	123	507.4	30.1	**0.0925**	0.032	34.2	ug/L	127	Standard
	Ba	135	38.7	5.4	**-0.0068**	0.001	14.6	ug/L	45	Standard
	Ce	140	40.0	21.7				ug/L	298	Standard
>	**Tb**	159	974680.3	0.5				ug/L	977942	Standard
	Ho	165	13.3	78.1				ug/L	12	Standard
	Tl	203	68.0	6.7	**0.0032**	0.001	15.5	ug/L	31	Standard
	Tl	205	66.7	11.5	**0.0068**	0.001	14.7	ug/L	20	Standard
	Pb	206	289.3	1.4	**-0.0104**	0.001	6.9	ug/L	286	Standard
	Pb	207	250.7	7.8	**-0.0070**	0.004	55.4	ug/L	257	Standard
	Pb	208	1099.7	2.8	**-0.0025**	0.002	60.9	ug/L	956	Standard
	U	238	62.3	42.5	**0.0057**	0.004	64.1	ug/L	17	Standard
>	**Bi**	209	560756.1	0.5				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:58:37
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	33.3	45.8	**-0.0130**	0.021	161.7	mg/L	32	Standard
K	39	33.3	17.3	**-0.0301**	0.018	60.0	mg/L	30	Standard
Ca	43	38.3	27.2	**0.4086**	1.513	370.3	mg/L	48	Standard
Fe	54	141.0	17.0	**-0.0107**	0.022	208.8	mg/L	159	Standard
Fe	57	213.3	13.7	**0.0887**	0.105	118.5	mg/L	195	Standard
> **Sc-1**	45	39966.0	1.7				mg/L	39543	Standard
Cl	35	34581.4	2.4				ug/L	27642	Standard
Kr	83	3.0	66.7				ug/L	2	Standard
Br	81	1706.8	4.2				ug/L	1647	Standard
P	31	18139.2	2.9				ug/L	15231	Standard
S	34	2418.5	8.0				ug/L	1845	Standard
Sr	88	68.3	11.2				ug/L	70	Standard
C	12	130.0	53.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.8	101.7				mg/L	12	Standard
Ho-1	165	13.3	78.1				mg/L	12	Standard
Er	166	3.3	173.2				mg/L	20	Standard
I	127	1121.7	18.5				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		98.559	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:58:37
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.089
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.660
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:58:37
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: PBW 45 WG573751-04
Sample Date/Time: Tuesday, June 28, 2016 15:59:33
Number of Replicates: 3
Autosampler Position: 232
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80427.9	0.7				ug/L	81760	Standard
	Be	9	20.0	43.3	**0.0107**	0.009	82.0	ug/L	17	Standard
	Al	27	16896.1	3.6	**0.0970**	0.004	4.0	ug/L	1410	Standard
	Sc	45	39630.1	2.3				ug/L	39543	Standard
	Ti	47	34.3	28.0	**-0.0290**	0.037	127.4	ug/L	39	Standard
	V	51	1241.7	7.6	**-0.0303**	0.012	39.4	ug/L	1415	Standard
	Cr	52	8808.2	2.5	**0.1923**	0.015	7.8	ug/L	6946	Standard
	Cr	53	705.0	11.9	**0.4526**	0.107	23.5	ug/L	328	Standard
	Mn	55	5471.7	0.3	**0.6851**	0.014	2.1	ug/L	823	Standard
	Co	59	272.7	7.3	**0.0070**	0.003	48.0	ug/L	252	Standard
	Ni	60	321.0	11.4	**0.0633**	0.019	30.2	ug/L	192	Standard
	Cu	65	232.7	11.6	**0.0301**	0.014	46.8	ug/L	141	Standard
	Zn	66	2018.1	1.4	**0.9905**	0.004	0.4	ug/L	491	Standard
>	**Ge**	72	521638.1	1.4				ug/L	515087	Standard
	As	75	-54.9	36.7	**-0.0079**	0.017	218.9	ug/L	20	Standard
	Se	82	10.0	20.3	**-0.0338**	0.018	54.1	ug/L	15	Standard
	Se-1	77	48.0	14.4	**-0.0888**	0.100	112.7	ug/L	48	Standard
>	**Ga**	71	53.3	28.6				mg/L	90	Standard
	Rb	85	156.7	21.7				ug/L	23	Standard
	Y	89	436462.5	0.8				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	82.2	21.1	**0.0108**	0.006	59.5	ug/L	11	Standard
	Ag	107	93.7	16.2	**-0.0025**	0.002	87.2	ug/L	134	Standard
	Cd	111	11.9	25.2	**0.0025**	0.001	54.5	mg/L	7	Standard
	Cd	114	46.6	47.9	**0.0039**	0.004	108.7	ug/L	16	Standard
>	**In**	115	622987.2	0.5				ug/L	611117	Standard
	Sn	118	873.4	7.7	**0.0429**	0.010	24.2	ug/L	518	Standard
	Sb	123	225.0	36.0	**0.0339**	0.017	49.6	ug/L	127	Standard
	Ba	135	668.7	0.9	**0.2926**	0.004	1.5	ug/L	45	Standard
	Ce	140	103.3	7.4				ug/L	298	Standard
>	**Tb**	159	980754.1	0.6				ug/L	977942	Standard
	Ho	165	13.3	78.1				ug/L	12	Standard
	Tl	203	31.7	30.7	**-0.0012**	0.001	94.5	ug/L	31	Standard
	Tl	205	30.0	86.6	**0.0017**	0.003	203.8	ug/L	20	Standard
	Pb	206	380.3	5.4	**0.0044**	0.003	72.4	ug/L	286	Standard
	Pb	207	331.3	8.5	**0.0076**	0.005	67.3	ug/L	257	Standard
	Pb	208	1302.7	2.5	**0.0063**	0.001	19.3	ug/L	956	Standard
	U	238	25.3	53.6	**0.0004**	0.002	408.2	ug/L	17	Standard
>	**Bi**	209	576453.2	0.7				ug/L	557080	Standard

Sample ID: PBW 45 WG573751-04
Report Date/Time: Tuesday, June 28, 2016 16:01:50
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	33.3	8.7	**-0.0129**	0.003	22.5	mg/L	32	Standard
K	39	48.3	15.8	**0.0218**	0.028	129.2	mg/L	30	Standard
Ca	43	18.3	78.7	**-2.5918**	2.118	81.7	mg/L	48	Standard
Fe	54	122.7	28.9	**-0.0266**	0.030	114.3	mg/L	159	Standard
Fe	57	223.3	24.1	**0.1269**	0.176	138.7	mg/L	195	Standard
Sc-1	45	39630.1	2.3				mg/L	39543	Standard
Cl	35	33807.0	1.7				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	1633.4	2.0				ug/L	1647	Standard
P	31	17535.1	2.2				ug/L	15231	Standard
S	34	2338.5	7.5				ug/L	1845	Standard
Sr	88	75.0	30.6				ug/L	70	Standard
C	12	106.7	19.5				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	9.4	3.0				mg/L	12	Standard
Ho-1	165	13.3	78.1				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	955.0	19.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		98.371	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		101.272	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: PBW 45 WG573751-04
Report Date/Time: Tuesday, June 28, 2016 16:01:50
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.942
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.478
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: LCSW 45 WG573451-05
Sample Date/Time: Tuesday, June 28, 2016 16:02:44
Number of Replicates: 3
Autosampler Position: 233
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81285.8	1.4				ug/L	81760	Standard
	Be	9	49444.0	4.0	**50.0648**	1.744	3.5	ug/L	17	Standard
	Al	27	14440.2	1.6	**0.0798**	0.002	2.1	ug/L	1410	Standard
	Sc	45	40522.5	4.2				ug/L	39543	Standard
	Ti	47	35.7	14.1	**-0.0217**	0.021	96.5	ug/L	39	Standard
	V	51	343445.8	1.5	**49.2175**	1.714	3.5	ug/L	1415	Standard
	Cr	52	355835.3	1.6	**50.0593**	1.026	2.0	ug/L	6946	Standard
	Cr	53	43403.8	4.1	**49.0831**	1.336	2.7	ug/L	328	Standard
	Mn	55	327263.1	1.9	**50.3439**	1.847	3.7	ug/L	823	Standard
	Co	59	357909.9	1.9	**50.9146**	1.589	3.1	ug/L	252	Standard
	Ni	60	94484.4	2.0	**50.0595**	0.992	2.0	ug/L	192	Standard
	Cu	65	95887.5	1.7	**50.3702**	0.997	2.0	ug/L	141	Standard
	Zn	66	56697.9	0.9	**50.8990**	0.801	1.6	ug/L	491	Standard
>	**Ge**	72	515232.3	2.4				ug/L	515087	Standard
	As	75	56137.8	1.3	**49.9742**	1.270	2.5	ug/L	20	Standard
	Se	82	5754.2	0.4	**49.8676**	1.316	2.6	ug/L	15	Standard
	Se-1	77	3847.2	2.4	**50.3547**	1.829	3.6	ug/L	48	Standard
>	**Ga**	71	80.0	38.0				mg/L	90	Standard
	Rb	85	45.0	58.8				ug/L	23	Standard
	Y	89	449389.0	2.2				ug/L	448510	Standard
>	**Rh**	103	28.3	66.8				ug/L	15	Standard
	Mo	98	72.0	12.6	**0.0070**	0.003	48.7	ug/L	11	Standard
	Ag	107	332360.6	2.0	**48.5009**	1.256	2.6	ug/L	134	Standard
	Cd	111	105950.1	2.1	**48.5677**	1.420	2.9	mg/L	7	Standard
	Cd	114	259244.9	2.6	**48.8950**	1.818	3.7	ug/L	16	Standard
>	**In**	115	623755.7	1.2				ug/L	611117	Standard
	Sn	118	950.0	6.3	**0.0554**	0.008	14.9	ug/L	518	Standard
	Sb	123	235787.4	1.4	**48.9583**	1.261	2.6	ug/L	127	Standard
	Ba	135	101234.2	2.5	**48.0272**	1.604	3.3	ug/L	45	Standard
	Ce	140	53.3	5.4				ug/L	298	Standard
>	**Tb**	159	962831.7	2.3				ug/L	977942	Standard
	Ho	165	6.7	114.6				ug/L	12	Standard
	Tl	203	424492.3	2.2	**49.7361**	1.545	3.1	ug/L	31	Standard
	Tl	205	378457.4	1.6	**51.0672**	1.520	3.0	ug/L	20	Standard
	Pb	206	276584.3	1.3	**50.4636**	1.380	2.7	ug/L	286	Standard
	Pb	207	241372.6	1.3	**48.8258**	1.366	2.8	ug/L	257	Standard
	Pb	208	990155.6	1.9	**51.7752**	1.712	3.3	ug/L	956	Standard
	U	238	358184.2	2.6	**49.3578**	1.903	3.9	ug/L	17	Standard
>	**Bi**	209	564381.2	1.4				ug/L	557080	Standard

Sample ID: LCSW 45 WG573451-05
Report Date/Time: Tuesday, June 28, 2016 16:05:01
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.8762**	18.829	173.1	mg/L	0	Standard
Mg	24	28.3	71.3	**-0.0199**	0.028	139.9	mg/L	32	Standard
K	39	45.0	40.1	**0.0061**	0.057	941.7	mg/L	30	Standard
Ca	43	25.0	60.0	**-1.6629**	2.211	133.0	mg/L	48	Standard
Fe	54	141.6	12.0	**-0.0123**	0.010	79.8	mg/L	159	Standard
Fe	57	180.0	5.6	**-0.0319**	0.009	26.7	mg/L	195	Standard
> **Sc-1**	45	40522.5	4.2				mg/L	39543	Standard
Cl	35	33285.9	3.1				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	5364.3	7.6				ug/L	1647	Standard
P	31	17471.7	4.0				ug/L	15231	Standard
S	34	2361.9	6.4				ug/L	1845	Standard
Sr	88	63.3	24.1				ug/L	70	Standard
C	12	126.7	43.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	15.9	34.8				mg/L	12	Standard
Ho-1	165	6.7	114.6				mg/L	12	Standard
Er	166	16.7	69.3				mg/L	20	Standard
I	127	856.7	31.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		99.420	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.028	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 45 WG573451-05
Report Date/Time: Tuesday, June 28, 2016 16:05:01
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.068
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.311
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115101 WG573751-02
Sample Date/Time: Tuesday, June 28, 2016 16:05:56
Number of Replicates: 3
Autosampler Position: 234
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77671.9	1.4				ug/L	81760	Standard
	Be	9	26.7	39.0	**0.0185**	0.011	58.4	ug/L	17	Standard
	Al	27	13955917.4	1.2	**95.3134**	2.493	2.6	ug/L	1410	Standard
	Sc	45	39321.0	0.9				ug/L	39543	Standard
	Ti	47	301.7	10.7	**1.0707**	0.148	13.8	ug/L	39	Standard
	V	51	16198.5	2.2	**2.1650**	0.025	1.2	ug/L	1415	Standard
	Cr	52	10911.9	4.6	**0.5376**	0.044	8.1	ug/L	6946	Standard
	Cr	53	3267.0	16.4	**3.4384**	0.551	16.0	ug/L	328	Standard
	Mn	55	181411.8	1.7	**28.3397**	0.987	3.5	ug/L	823	Standard
	Co	59	976.7	3.3	**0.1102**	0.006	5.4	ug/L	252	Standard
	Ni	60	1471.7	0.7	**0.6905**	0.022	3.2	ug/L	192	Standard
	Cu	65	1872.1	2.6	**0.9119**	0.023	2.5	ug/L	141	Standard
	Zn	66	6550.4	2.0	**5.2541**	0.063	1.2	ug/L	491	Standard
>	**Ge**	72	506178.8	2.0				ug/L	515087	Standard
	As	75	931.2	4.4	**0.8837**	0.051	5.8	ug/L	20	Standard
	Se	82	231.2	4.1	**1.9256**	0.108	5.6	ug/L	15	Standard
	Se-1	77	230.7	3.7	**2.3974**	0.119	5.0	ug/L	48	Standard
>	**Ga**	71	195.0	5.1				mg/L	90	Standard
	Rb	85	19537.6	1.2				ug/L	23	Standard
	Y	89	432314.6	1.1				ug/L	448510	Standard
>	**Rh**	103	61.7	12.4				ug/L	15	Standard
	Mo	98	3291.8	1.3	**1.2330**	0.013	1.0	ug/L	11	Standard
	Ag	107	150.7	8.2	**0.0069**	0.002	29.9	ug/L	134	Standard
	Cd	111	62.2	4.0	**0.0270**	0.001	5.4	mg/L	7	Standard
	Cd	114	153.0	45.9	**0.0254**	0.014	55.2	ug/L	16	Standard
>	**In**	115	594394.0	1.1				ug/L	611117	Standard
	Sn	118	883.4	9.5	**0.0516**	0.013	25.0	ug/L	518	Standard
	Sb	123	639.6	8.2	**0.1265**	0.011	8.6	ug/L	127	Standard
	Ba	135	68518.2	0.3	**34.1005**	0.460	1.3	ug/L	45	Standard
	Ce	140	2645.2	17.9				ug/L	298	Standard
>	**Tb**	159	950271.0	0.7				ug/L	977942	Standard
	Ho	165	53.3	10.8				ug/L	12	Standard
	Tl	203	411.7	14.6	**0.0459**	0.008	17.1	ug/L	31	Standard
	Tl	205	376.7	19.3	**0.0511**	0.011	20.8	ug/L	20	Standard
	Pb	206	1679.8	1.4	**0.2582**	0.003	1.3	ug/L	286	Standard
	Pb	207	1366.4	3.8	**0.2322**	0.009	3.7	ug/L	257	Standard
	Pb	208	5841.1	1.9	**0.2602**	0.004	1.4	ug/L	956	Standard
	U	238	9731.1	2.1	**1.4034**	0.022	1.5	ug/L	17	Standard
>	**Bi**	209	537950.1	0.8				ug/L	557080	Standard

Sample ID: L1606115101 WG573751-02
Report Date/Time: Tuesday, June 28, 2016 16:08:13
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	34736.4	1.4	**46.4420**	0.278	0.6	mg/L	32	Standard
K	39	261.7	6.7	**0.7474**	0.052	7.0	mg/L	30	Standard
Ca	43	178.3	24.5	**21.8549**	6.453	29.5	mg/L	48	Standard
Fe	54	262.3	8.0	**0.1018**	0.020	19.8	mg/L	159	Standard
Fe	57	341.7	2.2	**0.5353**	0.023	4.3	mg/L	195	Standard
Sc-1	45	39321.0	0.9				mg/L	39543	Standard
Cl	35	35275.1	3.8				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	35124.1	4.7				ug/L	1647	Standard
P	31	18164.2	1.2				ug/L	15231	Standard
S	34	2260.2	2.8				ug/L	1845	Standard
Sr	88	231.7	13.9				ug/L	70	Standard
C	12	256.7	41.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	63.6	33.9				mg/L	12	Standard
Ho-1	165	53.3	10.8				mg/L	12	Standard
Er	166	63.3	36.5				mg/L	20	Standard
I	127	311901.7	7.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		95.000	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.271	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115101 WG573751-02
Report Date/Time: Tuesday, June 28, 2016 16:08:13
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.264
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.566
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115103S WG573751-07
Sample Date/Time: Tuesday, June 28, 2016 16:09:07
Number of Replicates: 3
Autosampler Position: 235
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82738.6	0.2				ug/L	81760	Standard
	Be	9	50932.1	1.3	**50.6700**	0.765	1.5	ug/L	17	Standard
	Al	27	14256616.7	2.2	**91.3792**	1.822	2.0	ug/L	1410	Standard
	Sc	45	41140.8	1.3				ug/L	39543	Standard
	Ti	47	371.0	4.6	**1.3119**	0.075	5.7	ug/L	39	Standard
	V	51	355266.5	2.0	**50.3881**	1.016	2.0	ug/L	1415	Standard
	Cr	52	348734.5	0.9	**48.5348**	0.604	1.2	ug/L	6946	Standard
	Cr	53	45906.1	1.2	**51.4332**	1.330	2.6	ug/L	328	Standard
	Mn	55	484874.0	0.4	**73.8916**	1.228	1.7	ug/L	823	Standard
	Co	59	330526.6	1.2	**46.5427**	1.213	2.6	ug/L	252	Standard
	Ni	60	91208.4	1.4	**47.8307**	0.262	0.5	ug/L	192	Standard
	Cu	65	93139.1	0.3	**48.4331**	0.711	1.5	ug/L	141	Standard
	Zn	66	61775.3	0.7	**54.9696**	1.209	2.2	ug/L	491	Standard
>	**Ge**	72	520412.6	1.6				ug/L	515087	Standard
	As	75	57188.6	0.5	**50.3925**	0.724	1.4	ug/L	20	Standard
	Se	82	5643.2	0.1	**48.4036**	0.791	1.6	ug/L	15	Standard
	Se-1	77	3992.2	4.4	**51.7398**	2.452	4.7	ug/L	48	Standard
>	**Ga**	71	230.0	28.3				mg/L	90	Standard
	Rb	85	19777.9	1.1				ug/L	23	Standard
	Y	89	433897.5	1.4				ug/L	448510	Standard
>	**Rh**	103	81.7	39.4				ug/L	15	Standard
	Mo	98	3301.1	2.4	**1.2202**	0.040	3.3	ug/L	11	Standard
	Ag	107	315385.8	0.9	**47.6558**	0.862	1.8	ug/L	134	Standard
	Cd	111	99498.3	0.6	**47.2265**	0.833	1.8	mg/L	7	Standard
	Cd	114	242671.7	0.4	**47.3834**	0.530	1.1	ug/L	16	Standard
>	**In**	115	602380.0	1.2				ug/L	611117	Standard
	Sn	118	866.7	11.7	**0.0467**	0.016	33.5	ug/L	518	Standard
	Sb	123	233832.7	0.4	**50.2714**	0.769	1.5	ug/L	127	Standard
	Ba	135	168472.5	0.5	**82.7716**	1.205	1.5	ug/L	45	Standard
	Ce	140	3190.3	1.8				ug/L	298	Standard
>	**Tb**	159	975538.4	0.7				ug/L	977942	Standard
	Ho	165	55.0	50.6				ug/L	12	Standard
	Tl	203	405391.8	0.6	**48.6809**	0.381	0.8	ug/L	31	Standard
	Tl	205	349844.8	1.2	**48.3796**	0.623	1.3	ug/L	20	Standard
	Pb	206	268733.3	0.7	**50.2550**	0.909	1.8	ug/L	286	Standard
	Pb	207	232854.8	1.2	**48.2799**	1.141	2.4	ug/L	257	Standard
	Pb	208	923192.2	0.6	**49.4693**	0.366	0.7	ug/L	956	Standard
	U	238	355626.8	0.4	**50.2227**	0.621	1.2	ug/L	17	Standard
>	**Bi**	209	550570.6	1.2				ug/L	557080	Standard

Sample ID: L1606115103S WG573751-07
Report Date/Time: Tuesday, June 28, 2016 16:11:24
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	35867.4	1.9	**45.8327**	0.548	1.2	mg/L	32	Standard
K	39	318.3	2.4	**0.8927**	0.038	4.2	mg/L	30	Standard
Ca	43	141.7	5.4	**15.3452**	1.333	8.7	mg/L	48	Standard
Fe	54	233.9	13.6	**0.0666**	0.029	43.8	mg/L	159	Standard
Fe	57	373.3	5.1	**0.5873**	0.073	12.3	mg/L	195	Standard
> **Sc-1**	45	41140.8	1.3				mg/L	39543	Standard
Cl	35	36665.0	1.2				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	38074.5	2.7				ug/L	1647	Standard
P	31	19512.6	0.4				ug/L	15231	Standard
S	34	2426.9	4.9				ug/L	1845	Standard
Sr	88	241.7	11.4				ug/L	70	Standard
C	12	183.3	31.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	66.3	40.2				mg/L	12	Standard
Ho-1	165	55.0	50.6				mg/L	12	Standard
Er	166	76.7	15.1				mg/L	20	Standard
I	127	311628.4	3.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.197	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.034	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115103S WG573751-07
Report Date/Time: Tuesday, June 28, 2016 16:11:24
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.570
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.831
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115304SD WG573751-08
Sample Date/Time: Tuesday, June 28, 2016 16:12:19
Number of Replicates: 3
Autosampler Position: 236
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82914.6	1.7				ug/L	81760	Standard
	Be	9	51258.3	2.2	**50.8925**	1.250	2.5	ug/L	17	Standard
	Al	27	15399778.5	0.9	**98.5192**	1.923	2.0	ug/L	1410	Standard
	Sc	45	40084.7	2.9				ug/L	39543	Standard
	Ti	47	640.0	9.7	**2.4715**	0.230	9.3	ug/L	39	Standard
	V	51	358753.7	1.2	**52.7008**	0.680	1.3	ug/L	1415	Standard
	Cr	52	354784.4	1.1	**51.1849**	0.073	0.1	ug/L	6946	Standard
	Cr	53	47661.6	3.5	**55.3029**	1.433	2.6	ug/L	328	Standard
	Mn	55	529314.6	1.5	**83.5469**	1.579	1.9	ug/L	823	Standard
	Co	59	345503.9	1.1	**50.3773**	0.891	1.8	ug/L	252	Standard
	Ni	60	92308.7	1.0	**50.1355**	0.426	0.9	ug/L	192	Standard
	Cu	65	93613.0	1.4	**50.4100**	0.723	1.4	ug/L	141	Standard
	Zn	66	64567.0	0.6	**59.5553**	0.411	0.7	ug/L	491	Standard
>	**Ge**	72	502524.4	1.0				ug/L	515087	Standard
	As	75	56933.1	0.7	**51.9453**	0.162	0.3	ug/L	20	Standard
	Se	82	5947.1	1.7	**52.8275**	0.666	1.3	ug/L	15	Standard
	Se-1	77	3965.5	2.7	**53.2435**	1.801	3.4	ug/L	48	Standard
>	**Ga**	71	291.7	11.2				mg/L	90	Standard
	Rb	85	22576.8	1.0				ug/L	23	Standard
	Y	89	439067.8	2.2				ug/L	448510	Standard
>	**Rh**	103	73.3	7.9				ug/L	15	Standard
	Mo	98	3536.7	2.2	**1.2876**	0.021	1.6	ug/L	11	Standard
	Ag	107	318639.0	0.2	**47.3947**	0.683	1.4	ug/L	134	Standard
	Cd	111	102545.7	0.7	**47.9082**	0.269	0.6	mg/L	7	Standard
	Cd	114	249882.8	0.5	**48.0284**	0.370	0.8	ug/L	16	Standard
>	**In**	115	611940.4	1.2				ug/L	611117	Standard
	Sn	118	1045.0	4.1	**0.0745**	0.005	6.8	ug/L	518	Standard
	Sb	123	235512.7	0.7	**49.8406**	0.693	1.4	ug/L	127	Standard
	Ba	135	174610.3	1.9	**84.4342**	0.545	0.6	ug/L	45	Standard
	Ce	140	5954.5	2.4				ug/L	298	Standard
>	**Tb**	159	973344.6	0.2				ug/L	977942	Standard
	Ho	165	121.7	6.3				ug/L	12	Standard
	Tl	203	413217.7	0.6	**49.7797**	0.456	0.9	ug/L	31	Standard
	Tl	205	363310.0	1.7	**50.4036**	0.986	2.0	ug/L	20	Standard
	Pb	206	270026.6	0.5	**50.6539**	0.212	0.4	ug/L	286	Standard
	Pb	207	235774.0	0.4	**49.0358**	0.293	0.6	ug/L	257	Standard
	Pb	208	947837.4	0.9	**50.9548**	0.634	1.2	ug/L	956	Standard
	U	238	363550.4	1.8	**51.5051**	1.028	2.0	ug/L	17	Standard
>	**Bi**	209	548791.6	0.3				ug/L	557080	Standard

Sample ID: L1606115304SD WG573751-08
Report Date/Time: Tuesday, June 28, 2016 16:14:36
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.4900**	18.160	173.1	mg/L	0	Standard
Mg	24	38714.4	0.7	**50.8180**	1.880	3.7	mg/L	32	Standard
K	39	343.3	15.2	**1.0066**	0.203	20.2	mg/L	30	Standard
Ca	43	156.7	1.8	**18.1495**	1.124	6.2	mg/L	48	Standard
Fe	54	374.0	6.7	**0.1967**	0.015	7.7	mg/L	159	Standard
Fe	57	400.0	0.0	**0.7083**	0.040	5.6	mg/L	195	Standard
> **Sc-1**	45	40084.7	2.9				mg/L	39543	Standard
Cl	35	38416.6	0.8				ug/L	27642	Standard
Kr	83	0.3	173.2				ug/L	2	Standard
Br	81	40474.1	6.0				ug/L	1647	Standard
P	31	20383.7	1.9				ug/L	15231	Standard
S	34	2388.5	6.8				ug/L	1845	Standard
Sr	88	235.0	8.5				ug/L	70	Standard
C	12	253.3	25.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	175.2	19.1				mg/L	12	Standard
Ho-1	165	121.7	6.3				mg/L	12	Standard
Er	166	100.0	26.5				mg/L	20	Standard
I	127	354555.7	8.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.412	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.561	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115304SD WG573751-08
Report Date/Time: Tuesday, June 28, 2016 16:14:36
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	100.135	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	98.512	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091602 WG573751-01
Sample Date/Time: Tuesday, June 28, 2016 16:15:30
Number of Replicates: 3
Autosampler Position: 237
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83752.6	2.6				ug/L	81760	Standard
	Be	9	50.0	26.5	**0.0395**	0.014	34.7	ug/L	17	Standard
	Al	27	2701508.7	2.5	**17.0954**	0.130	0.8	ug/L	1410	Standard
	Sc	45	39755.5	3.1				ug/L	39543	Standard
	Ti	47	56.0	11.2	**0.0633**	0.027	42.8	ug/L	39	Standard
	V	51	1924.8	3.9	**0.0741**	0.009	12.6	ug/L	1415	Standard
	Cr	52	10285.8	1.8	**0.4408**	0.012	2.7	ug/L	6946	Standard
	Cr	53	1156.7	1.7	**0.9947**	0.041	4.2	ug/L	328	Standard
	Mn	55	253930.8	1.4	**39.5649**	0.851	2.1	ug/L	823	Standard
	Co	59	1951.8	2.9	**0.2502**	0.004	1.8	ug/L	252	Standard
	Ni	60	1859.1	3.7	**0.8956**	0.027	3.0	ug/L	192	Standard
	Cu	65	1000.0	2.6	**0.4428**	0.017	3.9	ug/L	141	Standard
	Zn	66	2280.8	1.9	**1.2819**	0.026	2.1	ug/L	491	Standard
>	**Ge**	72	508105.2	1.4				ug/L	515087	Standard
	As	75	151.5	17.1	**0.1768**	0.022	12.7	ug/L	20	Standard
	Se	82	39.2	12.0	**0.2256**	0.041	18.4	ug/L	15	Standard
	Se-1	77	70.3	2.2	**0.2277**	0.025	11.1	ug/L	48	Standard
>	**Ga**	71	73.3	3.9				mg/L	90	Standard
	Rb	85	9319.5	3.0				ug/L	23	Standard
	Y	89	446629.7	1.4				ug/L	448510	Standard
>	**Rh**	103	35.0	37.8				ug/L	15	Standard
	Mo	98	255.3	4.6	**0.0742**	0.003	4.3	ug/L	11	Standard
	Ag	107	140.0	17.3	**0.0044**	0.004	86.0	ug/L	134	Standard
	Cd	111	34.5	15.3	**0.0130**	0.002	17.2	mg/L	7	Standard
	Cd	114	78.1	40.8	**0.0099**	0.006	59.3	ug/L	16	Standard
>	**In**	115	618896.7	1.2				ug/L	611117	Standard
	Sn	118	878.4	9.4	**0.0447**	0.013	28.2	ug/L	518	Standard
	Sb	123	309.3	20.5	**0.0520**	0.014	26.9	ug/L	127	Standard
	Ba	135	35857.7	1.5	**17.1253**	0.158	0.9	ug/L	45	Standard
	Ce	140	1368.4	3.3				ug/L	298	Standard
>	**Tb**	159	968292.7	1.7				ug/L	977942	Standard
	Ho	165	100.0	15.0				ug/L	12	Standard
	Tl	203	175.7	13.5	**0.0159**	0.003	15.9	ug/L	31	Standard
	Tl	205	158.3	30.0	**0.0192**	0.006	31.5	ug/L	20	Standard
	Pb	206	469.3	2.7	**0.0227**	0.001	2.5	ug/L	286	Standard
	Pb	207	387.3	8.4	**0.0210**	0.006	30.5	ug/L	257	Standard
	Pb	208	1635.4	0.6	**0.0258**	0.001	5.1	ug/L	956	Standard
	U	238	220.3	6.6	**0.0276**	0.001	5.0	ug/L	17	Standard
>	**Bi**	209	560335.5	2.1				ug/L	557080	Standard

Sample ID: L1606091602 WG573751-01
Report Date/Time: Tuesday, June 28, 2016 16:17:47
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**20.5537**	35.591	173.2	mg/L	0	Standard
Mg	24	3380.4	0.3	**4.4213**	0.131	3.0	mg/L	32	Standard
K	39	168.3	9.1	**0.4237**	0.034	8.1	mg/L	30	Standard
Ca	43	108.3	14.8	**11.0098**	2.304	20.9	mg/L	48	Standard
Fe	54	282.4	10.4	**0.1178**	0.034	28.5	mg/L	159	Standard
Fe	57	291.7	21.2	**0.3561**	0.217	61.0	mg/L	195	Standard
> **Sc-1**	45	39755.5	3.1				mg/L	39543	Standard
Cl	35	40081.0	2.5				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	4644.1	13.8				ug/L	1647	Standard
P	31	18102.5	2.7				ug/L	15231	Standard
S	34	2331.8	2.8				ug/L	1845	Standard
Sr	88	118.3	8.8				ug/L	70	Standard
C	12	123.3	9.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	117.9	56.1				mg/L	12	Standard
Ho-1	165	100.0	15.0				mg/L	12	Standard
Er	166	43.3	48.0				mg/L	20	Standard
I	127	21574.4	25.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.437	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.645	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091602 WG573751-01
Report Date/Time: Tuesday, June 28, 2016 16:17:47
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.273
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.584
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091602DP WG573751-06
Sample Date/Time: Tuesday, June 28, 2016 16:18:41
Number of Replicates: 3
Autosampler Position: 238
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81912.6	2.5				ug/L	81760	Standard
	Be	9	38.3	32.8	**0.0286**	0.012	41.1	ug/L	17	Standard
	Al	27	2661315.3	2.5	**17.2245**	0.558	3.2	ug/L	1410	Standard
	Sc	45	39553.2	2.2				ug/L	39543	Standard
	Ti	47	60.3	10.8	**0.0810**	0.030	36.5	ug/L	39	Standard
	V	51	1771.3	1.7	**0.0516**	0.003	5.3	ug/L	1415	Standard
	Cr	52	10216.4	1.4	**0.4303**	0.018	4.2	ug/L	6946	Standard
	Cr	53	1328.4	4.4	**1.1923**	0.072	6.1	ug/L	328	Standard
	Mn	55	248102.9	0.7	**38.6399**	0.746	1.9	ug/L	823	Standard
	Co	59	1844.8	1.4	**0.2347**	0.001	0.3	ug/L	252	Standard
	Ni	60	1828.1	4.4	**0.8792**	0.056	6.4	ug/L	192	Standard
	Cu	65	995.0	2.4	**0.4399**	0.012	2.7	ug/L	141	Standard
	Zn	66	2165.5	0.6	**1.1746**	0.022	1.8	ug/L	491	Standard
>	**Ge**	72	508291.1	1.3				ug/L	515087	Standard
	As	75	90.4	12.4	**0.1217**	0.009	7.5	ug/L	20	Standard
	Se	82	34.1	11.8	**0.1803**	0.032	17.6	ug/L	15	Standard
	Se-1	77	74.3	14.5	**0.2819**	0.154	54.5	ug/L	48	Standard
>	**Ga**	71	68.3	30.5				mg/L	90	Standard
	Rb	85	9142.8	6.3				ug/L	23	Standard
	Y	89	450226.3	2.8				ug/L	448510	Standard
>	**Rh**	103	21.7	58.1				ug/L	15	Standard
	Mo	98	245.2	4.8	**0.0704**	0.005	7.4	ug/L	11	Standard
	Ag	107	87.7	6.3	**-0.0033**	0.001	21.6	ug/L	134	Standard
	Cd	111	23.2	17.5	**0.0078**	0.002	23.8	mg/L	7	Standard
	Cd	114	58.3	13.8	**0.0061**	0.001	23.6	ug/L	16	Standard
>	**In**	115	619551.5	1.3				ug/L	611117	Standard
	Sn	118	933.4	2.4	**0.0538**	0.006	10.7	ug/L	518	Standard
	Sb	123	201.2	19.7	**0.0293**	0.009	30.0	ug/L	127	Standard
	Ba	135	34661.9	0.9	**16.5369**	0.177	1.1	ug/L	45	Standard
	Ce	140	1260.1	4.6				ug/L	298	Standard
>	**Tb**	159	963340.8	1.2				ug/L	977942	Standard
	Ho	165	88.3	18.2				ug/L	12	Standard
	Tl	203	85.3	6.5	**0.0054**	0.001	13.0	ug/L	31	Standard
	Tl	205	83.3	17.3	**0.0091**	0.002	21.1	ug/L	20	Standard
	Pb	206	415.0	2.1	**0.0134**	0.001	8.9	ug/L	286	Standard
	Pb	207	359.0	2.4	**0.0158**	0.001	8.8	ug/L	257	Standard
	Pb	208	1485.0	2.9	**0.0185**	0.003	14.1	ug/L	956	Standard
	U	238	164.0	3.4	**0.0200**	0.001	3.7	ug/L	17	Standard
>	**Bi**	209	555403.6	0.6				ug/L	557080	Standard

Sample ID: L1606091602DP WG573751-06
Report Date/Time: Tuesday, June 28, 2016 16:20:58
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3443.7	5.9	**4.5314**	0.377	8.3	mg/L	32	Standard
K	39	158.3	7.9	**0.3934**	0.041	10.4	mg/L	30	Standard
Ca	43	93.3	12.4	**8.8061**	1.479	16.8	mg/L	48	Standard
Fe	54	304.0	6.5	**0.1380**	0.017	12.0	mg/L	159	Standard
Fe	57	313.3	6.4	**0.4340**	0.087	20.1	mg/L	195	Standard
> **Sc-1**	45	39553.2	2.2				mg/L	39543	Standard
Cl	35	40893.8	1.8				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	3857.2	4.5				ug/L	1647	Standard
P	31	18868.4	1.0				ug/L	15231	Standard
S	34	2383.5	7.0				ug/L	1845	Standard
Sr	88	95.0	13.9				ug/L	70	Standard
C	12	200.0	17.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	96.2	30.5				mg/L	12	Standard
Ho-1	165	88.3	18.2				mg/L	12	Standard
Er	166	80.0	37.5				mg/L	20	Standard
I	127	15653.1	0.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		100.187	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.681	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091602DP WG573751-06
Report Date/Time: Tuesday, June 28, 2016 16:20:58
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.380
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.699
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091601
Sample Date/Time: Tuesday, June 28, 2016 16:21:53
Number of Replicates: 3
Autosampler Position: 239
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81664.5	1.8				ug/L	81760	Standard
	Be	9	26.7	28.6	**0.0172**	0.008	47.1	ug/L	17	Standard
	Al	27	4275643.4	2.5	**27.7561**	0.463	1.7	ug/L	1410	Standard
	Sc	45	38087.8	3.0				ug/L	39543	Standard
	Ti	47	54.7	26.0	**0.0548**	0.059	108.3	ug/L	39	Standard
	V	51	1697.6	8.1	**0.0374**	0.021	56.8	ug/L	1415	Standard
	Cr	52	10296.2	0.8	**0.4201**	0.020	4.7	ug/L	6946	Standard
	Cr	53	1460.1	7.5	**1.3201**	0.135	10.2	ug/L	328	Standard
	Mn	55	64493.4	3.0	**9.7887**	0.389	4.0	ug/L	823	Standard
	Co	59	427.0	5.1	**0.0294**	0.004	12.3	ug/L	252	Standard
	Ni	60	847.0	2.5	**0.3443**	0.016	4.7	ug/L	192	Standard
	Cu	65	826.7	6.2	**0.3435**	0.022	6.5	ug/L	141	Standard
	Zn	66	9487.6	0.7	**7.8171**	0.052	0.7	ug/L	491	Standard
>	**Ge**	72	515789.8	1.2				ug/L	515087	Standard
	As	75	95.2	47.1	**0.1249**	0.040	32.1	ug/L	20	Standard
	Se	82	32.5	28.2	**0.1620**	0.079	48.7	ug/L	15	Standard
	Se-1	77	78.3	8.3	**0.3202**	0.098	30.6	ug/L	48	Standard
>	**Ga**	71	58.3	50.2				mg/L	90	Standard
	Rb	85	8367.3	1.9				ug/L	23	Standard
	Y	89	424791.4	2.4				ug/L	448510	Standard
>	**Rh**	103	81.7	19.7				ug/L	15	Standard
	Mo	98	367.2	2.9	**0.1192**	0.004	3.5	ug/L	11	Standard
	Ag	107	119.3	48.0	**0.0019**	0.009	460.3	ug/L	134	Standard
	Cd	111	50.7	68.6	**0.0212**	0.017	78.0	mg/L	7	Standard
	Cd	114	102.4	71.8	**0.0151**	0.014	95.2	ug/L	16	Standard
>	**In**	115	600349.6	1.0				ug/L	611117	Standard
	Sn	118	1276.7	2.5	**0.1178**	0.004	3.0	ug/L	518	Standard
	Sb	123	200.9	15.6	**0.0306**	0.007	23.5	ug/L	127	Standard
	Ba	135	168683.5	0.8	**83.1505**	0.843	1.0	ug/L	45	Standard
	Ce	140	160.0	17.4				ug/L	298	Standard
>	**Tb**	159	943339.1	0.8				ug/L	977942	Standard
	Ho	165	5.0	100.0				ug/L	12	Standard
	Tl	203	103.7	83.0	**0.0076**	0.010	135.3	ug/L	31	Standard
	Tl	205	65.0	38.5	**0.0066**	0.003	51.6	ug/L	20	Standard
	Pb	206	449.7	4.5	**0.0200**	0.004	18.4	ug/L	286	Standard
	Pb	207	373.7	5.2	**0.0190**	0.004	20.9	ug/L	257	Standard
	Pb	208	1469.0	1.7	**0.0178**	0.001	8.0	ug/L	956	Standard
	U	238	195.0	9.9	**0.0244**	0.003	11.1	ug/L	17	Standard
>	**Bi**	209	554105.5	0.2				ug/L	557080	Standard

Sample ID: L1606091601
Report Date/Time: Tuesday, June 28, 2016 16:24:10
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**11.2769**	19.523	173.1	mg/L	0	Standard
Mg	24	4872.5	3.1	**6.6831**	0.365	5.5	mg/L	32	Standard
K	39	180.0	11.1	**0.4891**	0.054	11.0	mg/L	30	Standard
Ca	43	71.7	49.5	**5.8566**	5.217	89.1	mg/L	48	Standard
Fe	54	187.0	32.7	**0.0391**	0.060	153.5	mg/L	159	Standard
Fe	57	275.0	13.7	**0.3422**	0.158	46.1	mg/L	195	Standard
> **Sc-1**	45	38087.8	3.0				mg/L	39543	Standard
Cl	35	39985.4	2.3				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	5007.5	9.4				ug/L	1647	Standard
P	31	18170.9	2.8				ug/L	15231	Standard
S	34	2413.5	5.6				ug/L	1845	Standard
Sr	88	318.3	14.3				ug/L	70	Standard
C	12	223.3	5.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	59.7	16.8				mg/L	12	Standard
Ho-1	165	5.0	100.0				mg/L	12	Standard
Er	166	6.7	86.6				mg/L	20	Standard
I	127	27700.4	6.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		99.883	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.136	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091601
Report Date/Time: Tuesday, June 28, 2016 16:24:10
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.238
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.466
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091601PS WG574522-01
Sample Date/Time: Tuesday, June 28, 2016 16:25:04
Number of Replicates: 3
Autosampler Position: 240
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84128.0	1.7				ug/L	81760	Standard
	Be	9	52619.6	0.9	**51.4982**	1.298	2.5	ug/L	17	Standard
	Al	27	4326633.2	0.4	**27.2707**	0.550	2.0	ug/L	1410	Standard
	Sc	45	39650.1	0.7				ug/L	39543	Standard
	Ti	47	62.3	14.0	**0.0872**	0.038	43.9	ug/L	39	Standard
	V	51	353435.0	1.4	**50.9243**	0.391	0.8	ug/L	1415	Standard
	Cr	52	362779.9	0.9	**51.3507**	0.534	1.0	ug/L	6946	Standard
	Cr	53	45635.3	3.2	**51.9244**	0.938	1.8	ug/L	328	Standard
	Mn	55	384138.6	0.3	**59.4407**	0.957	1.6	ug/L	823	Standard
	Co	59	352427.3	0.4	**50.4121**	0.778	1.5	ug/L	252	Standard
	Ni	60	95802.0	1.3	**51.0471**	0.630	1.2	ug/L	192	Standard
	Cu	65	98413.6	0.9	**51.9931**	0.640	1.2	ug/L	141	Standard
	Zn	66	63162.1	1.4	**57.1168**	0.348	0.6	ug/L	491	Standard
>	**Ge**	72	512293.3	1.9				ug/L	515087	Standard
	As	75	56566.0	1.5	**50.6330**	0.812	1.6	ug/L	20	Standard
	Se	82	5805.6	1.0	**50.5901**	0.677	1.3	ug/L	15	Standard
	Se-1	77	3921.2	1.4	**51.6223**	0.909	1.8	ug/L	48	Standard
>	**Ga**	71	60.0	50.7				mg/L	90	Standard
	Rb	85	8215.6	0.6				ug/L	23	Standard
	Y	89	438501.2	1.8				ug/L	448510	Standard
>	**Rh**	103	71.7	28.2				ug/L	15	Standard
	Mo	98	368.8	5.9	**0.1169**	0.008	6.7	ug/L	11	Standard
	Ag	107	334626.2	0.5	**49.6506**	0.360	0.7	ug/L	134	Standard
	Cd	111	105047.6	0.7	**48.9601**	0.529	1.1	mg/L	7	Standard
	Cd	114	254060.7	2.8	**48.7162**	1.522	3.1	ug/L	16	Standard
>	**In**	115	613389.4	0.3				ug/L	611117	Standard
	Sn	118	1936.8	37.5	**0.2241**	0.121	54.0	ug/L	518	Standard
	Sb	123	237358.9	1.9	**50.1080**	0.980	2.0	ug/L	127	Standard
	Ba	135	267735.0	1.3	**129.1838**	2.081	1.6	ug/L	45	Standard
	Ce	140	231.7	21.3				ug/L	298	Standard
>	**Tb**	159	964479.2	1.1				ug/L	977942	Standard
	Ho	165	18.3	56.8				ug/L	12	Standard
	Tl	203	432424.0	0.5	**50.0145**	0.262	0.5	ug/L	31	Standard
	Tl	205	382714.9	0.3	**50.9759**	0.183	0.4	ug/L	20	Standard
	Pb	206	282966.8	1.1	**50.9646**	0.635	1.2	ug/L	286	Standard
	Pb	207	246387.7	1.3	**49.1992**	0.690	1.4	ug/L	257	Standard
	Pb	208	985942.4	0.8	**50.8876**	0.399	0.8	ug/L	956	Standard
	U	238	365334.3	1.0	**49.6910**	0.475	1.0	ug/L	17	Standard
>	**Bi**	209	571593.5	0.1				ug/L	557080	Standard

Sample ID: L1606091601PS WG574522-01
Report Date/Time: Tuesday, June 28, 2016 16:27:21
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4890.8	3.1	**6.4355**	0.187	2.9	mg/L	32	Standard
K	39	161.7	23.2	**0.4031**	0.125	30.9	mg/L	30	Standard
Ca	43	113.3	24.3	**11.8087**	4.086	34.6	mg/L	48	Standard
Fe	54	177.2	7.7	**0.0228**	0.011	49.5	mg/L	159	Standard
Fe	57	255.0	15.6	**0.2338**	0.132	56.3	mg/L	195	Standard
> **Sc-1**	45	39650.1	0.7				mg/L	39543	Standard
Cl	35	40732.7	2.1				ug/L	27642	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	4914.1	11.2				ug/L	1647	Standard
P	31	18529.7	1.8				ug/L	15231	Standard
S	34	2476.9	6.8				ug/L	1845	Standard
Sr	88	310.0	4.8				ug/L	70	Standard
C	12	160.0	18.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	12.7	117.1				mg/L	12	Standard
Ho-1	165	18.3	56.8				mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	27421.5	4.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.896	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.458	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091601PS WG574522-01
Report Date/Time: Tuesday, June 28, 2016 16:27:21
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.372
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.605
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606091601PS WG574522-01
Report Date/Time: Tuesday, June 28, 2016 16:27:21
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091601SDL WG574522-02
Sample Date/Time: Tuesday, June 28, 2016 16:28:16
Number of Replicates: 3
Autosampler Position: 241
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82723.6	2.3				ug/L	81760	Standard
	Be	9	23.3	12.4	**0.0135**	0.003	18.7	ug/L	17	Standard
	Al	27	900313.9	2.1	**5.7596**	0.136	2.4	ug/L	1410	Standard
	Sc	45	39763.8	0.8				ug/L	39543	Standard
	Ti	47	29.7	9.7	**-0.0430**	0.012	28.1	ug/L	39	Standard
	V	51	1735.9	7.6	**0.0491**	0.020	40.6	ug/L	1415	Standard
	Cr	52	8352.3	2.2	**0.1715**	0.035	20.7	ug/L	6946	Standard
	Cr	53	645.0	2.8	**0.4109**	0.024	5.9	ug/L	328	Standard
	Mn	55	14377.5	0.8	**2.1221**	0.018	0.8	ug/L	823	Standard
	Co	59	301.0	11.3	**0.0125**	0.005	41.5	ug/L	252	Standard
	Ni	60	410.7	1.2	**0.1183**	0.004	3.1	ug/L	192	Standard
	Cu	65	611.3	5.6	**0.2385**	0.017	7.2	ug/L	141	Standard
	Zn	66	7975.8	0.8	**6.6226**	0.085	1.3	ug/L	491	Standard
>	**Ge**	72	503109.1	0.7				ug/L	515087	Standard
	As	75	19.7	109.3	**0.0582**	0.020	33.7	ug/L	20	Standard
	Se	82	17.8	58.3	**0.0388**	0.092	237.9	ug/L	15	Standard
	Se-1	77	57.7	14.7	**0.0653**	0.120	183.5	ug/L	48	Standard
>	**Ga**	71	58.3	44.0				mg/L	90	Standard
	Rb	85	1753.4	3.2				ug/L	23	Standard
	Y	89	444634.9	0.6				ug/L	448510	Standard
>	**Rh**	103	21.7	48.0				ug/L	15	Standard
	Mo	98	88.5	3.1	**0.0131**	0.001	5.4	ug/L	11	Standard
	Ag	107	142.0	4.2	**0.0046**	0.001	16.3	ug/L	134	Standard
	Cd	111	23.2	23.8	**0.0077**	0.002	31.6	mg/L	7	Standard
	Cd	114	28.0	34.6	**0.0004**	0.002	477.3	ug/L	16	Standard
>	**In**	115	620921.1	1.0				ug/L	611117	Standard
	Sn	118	821.7	5.2	**0.0348**	0.006	16.8	ug/L	518	Standard
	Sb	123	2051.2	21.4	**0.4156**	0.096	23.0	ug/L	127	Standard
	Ba	135	34796.2	0.2	**16.5640**	0.119	0.7	ug/L	45	Standard
	Ce	140	63.3	12.1				ug/L	298	Standard
>	**Tb**	159	973024.2	1.8				ug/L	977942	Standard
	Ho	165	8.3	124.9				ug/L	12	Standard
	Tl	203	97.0	5.7	**0.0066**	0.000	6.5	ug/L	31	Standard
	Tl	205	68.3	44.1	**0.0070**	0.004	57.1	ug/L	20	Standard
	Pb	206	360.7	11.0	**0.0024**	0.006	236.3	ug/L	286	Standard
	Pb	207	322.3	9.7	**0.0075**	0.006	86.2	ug/L	257	Standard
	Pb	208	1339.4	1.1	**0.0099**	0.001	10.7	ug/L	956	Standard
	U	238	86.3	6.6	**0.0090**	0.001	5.9	ug/L	17	Standard
>	**Bi**	209	562189.7	2.5				ug/L	557080	Standard

Sample ID: L1606091601SDL WG574522-02
Report Date/Time: Tuesday, June 28, 2016 16:30:33
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	991.7	7.3	**1.2554**	0.092	7.3	mg/L	32	Standard
K	39	51.7	11.2	**0.0323**	0.021	64.1	mg/L	30	Standard
Ca	43	40.0	45.1	**0.6884**	2.694	391.4	mg/L	48	Standard
Fe	54	140.9	19.7	**-0.0104**	0.024	229.0	mg/L	159	Standard
Fe	57	241.7	4.8	**0.1865**	0.032	17.2	mg/L	195	Standard
> **Sc-1**	45	39763.8	0.8				mg/L	39543	Standard
Cl	35	40635.1	0.7				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	2310.2	7.5				ug/L	1647	Standard
P	31	19332.3	4.6				ug/L	15231	Standard
S	34	2451.9	6.9				ug/L	1845	Standard
Sr	88	125.0	21.2				ug/L	70	Standard
C	12	163.3	25.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	19.5	88.8				mg/L	12	Standard
Ho-1	165	8.3	124.9				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	7598.6	6.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.179	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.675	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091601SDL WG574522-02
Report Date/Time: Tuesday, June 28, 2016 16:30:33
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.604
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.917
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 16:31:29
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	79881.8	2.3				ug/L	81760	Standard
	Be	9	50840.2	1.3	**52.4015**	1.145	2.2	ug/L	17	Standard
	Al	27	7585830.9	1.9	**50.3677**	1.194	2.4	ug/L	1410	Standard
	Sc	45	39128.8	0.6				ug/L	39543	Standard
	Ti	47	24998.6	0.9	**102.1788**	1.762	1.7	ug/L	39	Standard
	V	51	348429.9	1.4	**50.8229**	0.271	0.5	ug/L	1415	Standard
	Cr	52	355416.3	0.4	**50.9232**	0.579	1.1	ug/L	6946	Standard
	Cr	53	44443.4	1.1	**51.2031**	0.906	1.8	ug/L	328	Standard
	Mn	55	315927.2	0.8	**49.4624**	0.575	1.2	ug/L	823	Standard
	Co	59	353189.3	0.9	**51.1457**	0.854	1.7	ug/L	252	Standard
	Ni	60	94600.8	0.9	**51.0293**	0.136	0.3	ug/L	192	Standard
	Cu	65	95833.5	0.4	**51.2541**	0.406	0.8	ug/L	141	Standard
	Zn	66	56109.7	0.5	**51.2906**	0.852	1.7	ug/L	491	Standard
>	Ge	72	505999.1	1.2				ug/L	515087	Standard
	As	75	55945.5	0.4	**50.6968**	0.402	0.8	ug/L	20	Standard
	Se	82	5756.3	0.8	**50.7852**	1.001	2.0	ug/L	15	Standard
	Se-1	77	3889.8	1.1	**51.8484**	1.021	2.0	ug/L	48	Standard
>	Ga	71	51.7	20.1				mg/L	90	Standard
	Rb	85	811.7	9.0				ug/L	23	Standard
	Y	89	426423.1	2.5				ug/L	448510	Standard
>	Rh	103	31.7	24.1				ug/L	15	Standard
	Mo	98	265386.1	0.2	**99.3199**	0.619	0.6	ug/L	11	Standard
	Ag	107	333927.0	1.2	**50.3137**	0.313	0.6	ug/L	134	Standard
	Cd	111	105484.6	1.1	**49.9277**	0.764	1.5	mg/L	7	Standard
	Cd	114	258596.4	2.6	**50.3494**	1.108	2.2	ug/L	16	Standard
>	In	115	604027.1	0.8				ug/L	611117	Standard
	Sn	118	295796.4	1.2	**50.5150**	0.923	1.8	ug/L	518	Standard
	Sb	123	237014.9	1.1	**50.8107**	0.333	0.7	ug/L	127	Standard
	Ba	135	100914.0	1.4	**49.4318**	0.852	1.7	ug/L	45	Standard
	Ce	140	80.0	0.0				ug/L	298	Standard
>	Tb	159	968181.4	1.1				ug/L	977942	Standard
	Ho	165	8.3	69.3				ug/L	12	Standard
	Tl	203	430250.2	1.2	**50.7804**	1.115	2.2	ug/L	31	Standard
	Tl	205	381434.8	2.3	**51.8414**	1.366	2.6	ug/L	20	Standard
	Pb	206	273103.8	0.8	**50.1867**	0.395	0.8	ug/L	286	Standard
	Pb	207	246704.6	0.5	**50.2667**	0.621	1.2	ug/L	257	Standard
	Pb	208	974193.4	0.6	**51.3084**	0.901	1.8	ug/L	956	Standard
	U	238	366221.1	0.4	**50.8280**	0.746	1.5	ug/L	17	Standard
>	Bi	209	560252.8	1.6				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 16:33:46
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3943.8	5.4	**5.2490**	0.301	5.7	mg/L	32	Standard
K	39	1356.7	5.3	**4.4904**	0.223	5.0	mg/L	30	Standard
Ca	43	83.3	18.3	**7.4389**	2.295	30.8	mg/L	48	Standard
Fe	54	5957.3	2.0	**5.3112**	0.093	1.8	mg/L	159	Standard
Fe	57	1600.1	10.0	**4.8365**	0.580	12.0	mg/L	195	Standard
> **Sc-1**	45	39128.8	0.6				mg/L	39543	Standard
Cl	35	37049.3	3.3				ug/L	27642	Standard
Kr	83	2.0	132.3				ug/L	2	Standard
Br	81	1520.1	14.1				ug/L	1647	Standard
P	31	19898.1	2.1				ug/L	15231	Standard
S	34	2718.6	2.9				ug/L	1845	Standard
Sr	88	73.3	20.8				ug/L	70	Standard
C	12	120.0	43.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	9.0	113.3				mg/L	12	Standard
Ho-1	165	8.3	69.3				mg/L	12	Standard
Er	166	20.0	50.0				mg/L	20	Standard
I	127	2098.5	16.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	104.803		
Al	27	100.735		
Sc	45			
Ti	47	102.179		
V	51	101.646		
Cr	52	101.846		
Cr	53			
Mn	55	98.925		
Co	59	102.291		
Ni	60	102.059		
Cu	65	102.508		
Zn	66	102.581		
> Ge	72		98.236	
As	75	101.394		
Se	82	101.570		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 16:33:46
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	99.320		
Ag	107	100.627		
Cd	111	99.855		
Cd	114			
> In	115			98.840
Sn	118	101.030		
Sb	123	101.621		
Ba	135	98.864		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.561		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.617		
U	238	101.656		
> Bi	209			100.569
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 16:33:46
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 16:34:41
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77653.6	3.2				ug/L	81760	Standard
	Be	9	28.3	27.0	**0.0205**	0.009	43.8	ug/L	17	Standard
	Al	27	2033.5	6.8	**-0.0005**	0.001	113.9	ug/L	1410	Standard
	Sc	45	38149.6	0.7				ug/L	39543	Standard
	Ti	47	33.0	16.0	**-0.0290**	0.023	78.1	ug/L	39	Standard
	V	51	1124.7	5.5	**-0.0407**	0.010	23.8	ug/L	1415	Standard
	Cr	52	6085.9	0.6	**-0.1604**	0.016	10.1	ug/L	6946	Standard
	Cr	53	345.0	11.5	**0.0612**	0.041	67.1	ug/L	328	Standard
	Mn	55	652.0	2.2	**-0.0453**	0.002	5.1	ug/L	823	Standard
	Co	59	236.0	7.4	**0.0031**	0.003	92.5	ug/L	252	Standard
	Ni	60	163.3	9.5	**-0.0160**	0.008	51.9	ug/L	192	Standard
	Cu	65	132.3	11.7	**-0.0193**	0.009	47.2	ug/L	141	Standard
	Zn	66	446.3	5.3	**-0.4102**	0.017	4.3	ug/L	491	Standard
>	**Ge**	72	502289.8	1.3				ug/L	515087	Standard
	As	75	-42.3	118.5	**0.0014**	0.046	3304.8	ug/L	20	Standard
	Se	82	15.3	48.7	**0.0168**	0.064	382.8	ug/L	15	Standard
	Se-1	77	51.0	21.6	**-0.0257**	0.142	551.0	ug/L	48	Standard
>	**Ga**	71	38.3	27.2				mg/L	90	Standard
	Rb	85	10.0	86.6				ug/L	23	Standard
	Y	89	438316.4	0.6				ug/L	448510	Standard
>	**Rh**	103	15.0	0.0				ug/L	15	Standard
	Mo	98	228.3	18.4	**0.0667**	0.015	22.0	ug/L	11	Standard
	Ag	107	123.7	9.2	**0.0025**	0.002	72.5	ug/L	134	Standard
	Cd	111	11.3	56.7	**0.0024**	0.003	123.9	mg/L	7	Standard
	Cd	114	32.3	39.0	**0.0014**	0.003	178.7	ug/L	16	Standard
>	**In**	115	601315.6	1.3				ug/L	611117	Standard
	Sn	118	783.4	33.2	**0.0331**	0.047	140.5	ug/L	518	Standard
	Sb	123	646.2	31.1	**0.1267**	0.045	35.7	ug/L	127	Standard
	Ba	135	40.3	7.6	**-0.0052**	0.002	33.4	ug/L	45	Standard
	Ce	140	38.3	19.9				ug/L	298	Standard
>	**Tb**	159	937082.2	1.4				ug/L	977942	Standard
	Ho	165	5.0	100.0				ug/L	12	Standard
	Tl	203	49.7	12.9	**0.0012**	0.001	66.7	ug/L	31	Standard
	Tl	205	40.0	57.3	**0.0033**	0.003	97.6	ug/L	20	Standard
	Pb	206	275.3	5.0	**-0.0116**	0.003	25.8	ug/L	286	Standard
	Pb	207	258.0	7.0	**-0.0041**	0.003	84.6	ug/L	257	Standard
	Pb	208	1051.3	3.4	**-0.0036**	0.002	41.9	ug/L	956	Standard
	U	238	50.3	16.2	**0.0042**	0.001	28.1	ug/L	17	Standard
>	**Bi**	209	546028.7	0.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 16:36:58
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**22.1550**	19.184	86.6	mg/L	0	Standard
Mg	24	31.7	50.8	**-0.0134**	0.022	167.6	mg/L	32	Standard
K	39	30.0	28.9	**-0.0362**	0.031	85.5	mg/L	30	Standard
Ca	43	30.0	66.7	**-0.6084**	3.177	522.1	mg/L	48	Standard
Fe	54	168.0	24.0	**0.0206**	0.039	187.8	mg/L	159	Standard
Fe	57	143.3	24.5	**-0.1229**	0.124	101.2	mg/L	195	Standard
Sc-1	45	38149.6	0.7				mg/L	39543	Standard
Cl	35	36673.0	2.0				ug/L	27642	Standard
Kr	83	1.0	173.2				ug/L	2	Standard
Br	81	1570.1	7.0				ug/L	1647	Standard
P	31	19050.3	3.5				ug/L	15231	Standard
S	34	2736.9	0.8				ug/L	1845	Standard
Sr	88	48.3	15.8				ug/L	70	Standard
C	12	133.3	4.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	5.7	217.0				mg/L	12	Standard
Ho-1	165	5.0	100.0				mg/L	12	Standard
Er	166	20.0	100.0				mg/L	20	Standard
I	127	1683.4	18.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		97.516	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 16:36:58
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.396
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.016
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 16:36:58
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091601SDL WG574522-02
Sample Date/Time: Tuesday, June 28, 2016 16:37:54
Number of Replicates: 3
Autosampler Position: 242
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80481.5	1.5				ug/L	81760	Standard
	Be	9	31.7	9.1	**0.0226**	0.003	11.9	ug/L	17	Standard
	Al	27	181010.0	1.7	**1.1789**	0.039	3.3	ug/L	1410	Standard
	Sc	45	40285.2	2.7				ug/L	39543	Standard
	Ti	47	30.3	10.1	**-0.0420**	0.014	34.3	ug/L	39	Standard
	V	51	1262.3	9.3	**-0.0236**	0.014	61.0	ug/L	1415	Standard
	Cr	52	6637.8	2.8	**-0.0958**	0.014	14.1	ug/L	6946	Standard
	Cr	53	403.3	17.4	**0.1211**	0.074	61.5	ug/L	328	Standard
	Mn	55	3292.4	3.2	**0.3636**	0.014	3.8	ug/L	823	Standard
	Co	59	275.7	13.5	**0.0082**	0.005	64.5	ug/L	252	Standard
	Ni	60	218.7	22.1	**0.0120**	0.025	207.8	ug/L	192	Standard
	Cu	65	189.7	2.7	**0.0098**	0.003	33.4	ug/L	141	Standard
	Zn	66	2022.1	5.3	**1.0328**	0.100	9.7	ug/L	491	Standard
>	**Ge**	72	510876.9	1.9				ug/L	515087	Standard
	As	75	-16.1	411.9	**0.0255**	0.059	233.0	ug/L	20	Standard
	Se	82	17.9	79.1	**0.0373**	0.123	328.7	ug/L	15	Standard
	Se-1	77	53.0	5.0	**-0.0093**	0.040	429.4	ug/L	48	Standard
>	**Ga**	71	26.7	10.8				mg/L	90	Standard
	Rb	85	385.0	11.5				ug/L	23	Standard
	Y	89	434970.9	1.3				ug/L	448510	Standard
>	**Rh**	103	13.3	57.3				ug/L	15	Standard
	Mo	98	105.1	23.0	**0.0193**	0.009	45.6	ug/L	11	Standard
	Ag	107	160.3	33.6	**0.0074**	0.008	107.2	ug/L	134	Standard
	Cd	111	27.8	77.7	**0.0099**	0.010	100.3	mg/L	7	Standard
	Cd	114	89.4	76.6	**0.0121**	0.013	107.6	ug/L	16	Standard
>	**In**	115	618555.8	0.1				ug/L	611117	Standard
	Sn	118	705.0	14.2	**0.0159**	0.017	105.2	ug/L	518	Standard
	Sb	123	323.1	28.2	**0.0548**	0.019	34.7	ug/L	127	Standard
	Ba	135	6967.6	1.3	**3.3092**	0.040	1.2	ug/L	45	Standard
	Ce	140	23.3	65.5				ug/L	298	Standard
>	**Tb**	159	976647.8	1.4				ug/L	977942	Standard
	Ho	165	15.0	57.7				ug/L	12	Standard
	Tl	203	112.7	67.1	**0.0081**	0.009	105.4	ug/L	31	Standard
	Tl	205	90.0	29.4	**0.0096**	0.003	35.2	ug/L	20	Standard
	Pb	206	345.7	14.3	**-0.0019**	0.008	437.8	ug/L	286	Standard
	Pb	207	297.3	11.3	**0.0008**	0.006	736.1	ug/L	257	Standard
	Pb	208	1206.4	8.6	**0.0013**	0.005	370.6	ug/L	956	Standard
	U	238	79.3	55.4	**0.0077**	0.006	75.6	ug/L	17	Standard
>	**Bi**	209	576852.7	1.1				ug/L	557080	Standard

Sample ID: L1606091601SDL WG574522-02
Report Date/Time: Tuesday, June 28, 2016 16:40:10
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	251.7	21.2	**0.2729**	0.077	28.4	mg/L	32	Standard
K	39	35.0	75.6	**-0.0238**	0.089	375.8	mg/L	30	Standard
Ca	43	41.7	66.1	**0.8623**	4.157	482.1	mg/L	48	Standard
Fe	54	139.7	6.2	**-0.0128**	0.010	79.6	mg/L	159	Standard
Fe	57	215.0	8.1	**0.0878**	0.054	61.5	mg/L	195	Standard
Sc-1	45	40285.2	2.7				mg/L	39543	Standard
Cl	35	36209.9	2.0				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1783.4	3.4				ug/L	1647	Standard
P	31	18659.8	1.9				ug/L	15231	Standard
S	34	2565.2	1.9				ug/L	1845	Standard
Sr	88	76.7	26.4				ug/L	70	Standard
C	12	126.7	24.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	19.2	52.1				mg/L	12	Standard
Ho-1	165	15.0	57.7				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	2670.2	2.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		98.436	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.183	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606091601SDL WG574522-02
Report Date/Time: Tuesday, June 28, 2016 16:40:10
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.217
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.549
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606114201
Sample Date/Time: Tuesday, June 28, 2016 16:41:05
Number of Replicates: 3
Autosampler Position: 243
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80520.1	1.9				ug/L	81760	Standard
	Be	9	30.0	16.7	**0.0209**	0.005	21.7	ug/L	17	Standard
	Al	27	1224587.7	1.8	**8.0560**	0.286	3.6	ug/L	1410	Standard
	Sc	45	42140.2	1.9				ug/L	39543	Standard
	Ti	47	397.3	21.3	**1.4906**	0.359	24.1	ug/L	39	Standard
	V	51	2122.9	0.3	**0.1097**	0.002	1.5	ug/L	1415	Standard
	Cr	52	14472.9	1.3	**1.0963**	0.024	2.2	ug/L	6946	Standard
	Cr	53	2878.6	1.5	**3.0550**	0.050	1.6	ug/L	328	Standard
	Mn	55	698464.2	0.4	**111.4396**	0.282	0.3	ug/L	823	Standard
	Co	59	22132.5	0.4	**3.2312**	0.012	0.4	ug/L	252	Standard
	Ni	60	7435.8	1.3	**3.9843**	0.043	1.1	ug/L	192	Standard
	Cu	65	724.4	2.6	**0.3040**	0.008	2.7	ug/L	141	Standard
	Zn	66	6121.2	1.0	**4.9565**	0.034	0.7	ug/L	491	Standard
>	**Ge**	72	497323.9	0.5				ug/L	515087	Standard
	As	75	171.3	19.9	**0.1980**	0.031	15.4	ug/L	20	Standard
	Se	82	64.0	10.3	**0.4560**	0.056	12.4	ug/L	15	Standard
	Se-1	77	106.7	7.0	**0.7479**	0.110	14.7	ug/L	48	Standard
>	**Ga**	71	88.3	6.5				mg/L	90	Standard
	Rb	85	4614.0	1.5				ug/L	23	Standard
	Y	89	442742.2	2.0				ug/L	448510	Standard
>	**Rh**	103	21.7	35.3				ug/L	15	Standard
	Mo	98	335.3	5.4	**0.1033**	0.008	7.8	ug/L	11	Standard
	Ag	107	92.0	14.2	**-0.0027**	0.002	71.5	ug/L	134	Standard
	Cd	111	86.4	7.6	**0.0370**	0.004	9.6	mg/L	7	Standard
	Cd	114	204.2	9.6	**0.0339**	0.004	12.0	ug/L	16	Standard
>	**In**	115	619678.1	1.4				ug/L	611117	Standard
	Sn	118	898.4	6.0	**0.0480**	0.011	22.7	ug/L	518	Standard
	Sb	123	324.0	26.0	**0.0551**	0.019	33.6	ug/L	127	Standard
	Ba	135	76873.5	0.9	**36.7017**	0.677	1.8	ug/L	45	Standard
	Ce	140	3832.2	0.5				ug/L	298	Standard
>	**Tb**	159	963958.5	1.3				ug/L	977942	Standard
	Ho	165	96.7	23.3				ug/L	12	Standard
	Tl	203	122.3	9.5	**0.0098**	0.001	14.0	ug/L	31	Standard
	Tl	205	126.7	13.9	**0.0151**	0.002	14.9	ug/L	20	Standard
	Pb	206	564.7	2.8	**0.0417**	0.003	6.6	ug/L	286	Standard
	Pb	207	477.0	3.5	**0.0406**	0.003	8.6	ug/L	257	Standard
	Pb	208	2018.4	2.1	**0.0474**	0.002	3.2	ug/L	956	Standard
	U	238	92.7	4.5	**0.0101**	0.001	6.9	ug/L	17	Standard
>	**Bi**	209	552929.6	1.0				ug/L	557080	Standard

Sample ID: L1606114201
Report Date/Time: Tuesday, June 28, 2016 16:43:22
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.1351**	17.546	173.1	mg/L	0	Standard
Mg	24	7430.2	1.6	**9.2261**	0.222	2.4	mg/L	32	Standard
K	39	131.7	22.3	**0.2750**	0.084	30.6	mg/L	30	Standard
Ca	43	50.0	10.0	**1.7747**	0.605	34.1	mg/L	48	Standard
Fe	54	2749.1	1.6	**2.1977**	0.008	0.4	mg/L	159	Standard
Fe	57	918.4	0.8	**2.2851**	0.038	1.7	mg/L	195	Standard
> **Sc-1**	45	42140.2	1.9				mg/L	39543	Standard
Cl	35	36101.6	0.8				ug/L	27642	Standard
Kr	83	3.3	75.5				ug/L	2	Standard
Br	81	12138.2	7.4				ug/L	1647	Standard
P	31	25736.8	1.2				ug/L	15231	Standard
S	34	2443.5	6.8				ug/L	1845	Standard
Sr	88	88.3	31.2				ug/L	70	Standard
C	12	360.0	10.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	87.0	21.3				mg/L	12	Standard
Ho-1	165	96.7	23.3				mg/L	12	Standard
Er	166	63.3	48.2				mg/L	20	Standard
I	127	223077.6	8.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		98.483	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.551	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606114201
Report Date/Time: Tuesday, June 28, 2016 16:43:22
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	101.401	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	99.255	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606114201
Report Date/Time: Tuesday, June 28, 2016 16:43:22
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606114203
Sample Date/Time: Tuesday, June 28, 2016 16:44:17
Number of Replicates: 3
Autosampler Position: 244
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81992.9	1.1				ug/L	81760	Standard
	Be	9	28.3	27.0	**0.0187**	0.008	41.6	ug/L	17	Standard
	Al	27	794059.8	1.7	**5.1234**	0.139	2.7	ug/L	1410	Standard
	Sc	45	39785.5	1.2				ug/L	39543	Standard
	Ti	47	377.0	7.0	**1.3367**	0.105	7.9	ug/L	39	Standard
	V	51	6234.1	1.0	**0.6819**	0.005	0.8	ug/L	1415	Standard
	Cr	52	12611.2	0.3	**0.7377**	0.006	0.8	ug/L	6946	Standard
	Cr	53	1423.4	7.9	**1.2652**	0.132	10.4	ug/L	328	Standard
	Mn	55	579693.3	1.0	**88.4369**	1.228	1.4	ug/L	823	Standard
	Co	59	3503.7	1.5	**0.4627**	0.007	1.6	ug/L	252	Standard
	Ni	60	3465.4	2.9	**1.7179**	0.054	3.1	ug/L	192	Standard
	Cu	65	1070.0	1.6	**0.4670**	0.008	1.6	ug/L	141	Standard
	Zn	66	3004.0	2.3	**1.8876**	0.071	3.8	ug/L	491	Standard
>	**Ge**	72	519952.2	0.4				ug/L	515087	Standard
	As	75	346.7	9.3	**0.3458**	0.028	8.2	ug/L	20	Standard
	Se	82	27.6	23.7	**0.1179**	0.056	47.6	ug/L	15	Standard
	Se-1	77	61.3	6.2	**0.0875**	0.047	53.2	ug/L	48	Standard
>	**Ga**	71	168.3	22.5				mg/L	90	Standard
	Rb	85	8202.2	4.2				ug/L	23	Standard
	Y	89	450084.2	2.9				ug/L	448510	Standard
>	**Rh**	103	20.0	50.0				ug/L	15	Standard
	Mo	98	804.4	3.7	**0.2777**	0.016	5.8	ug/L	11	Standard
	Ag	107	104.7	4.0	**-0.0007**	0.001	91.7	ug/L	134	Standard
	Cd	111	177.3	6.3	**0.0797**	0.005	6.4	mg/L	7	Standard
	Cd	114	412.5	12.7	**0.0743**	0.011	14.2	ug/L	16	Standard
>	**In**	115	613204.5	2.4				ug/L	611117	Standard
	Sn	118	838.4	3.1	**0.0394**	0.004	11.4	ug/L	518	Standard
	Sb	123	255.2	2.7	**0.0411**	0.003	6.6	ug/L	127	Standard
	Ba	135	46232.8	2.2	**22.2974**	0.510	2.3	ug/L	45	Standard
	Ce	140	10413.6	4.0				ug/L	298	Standard
>	**Tb**	159	959024.8	2.1				ug/L	977942	Standard
	Ho	165	235.0	14.0				ug/L	12	Standard
	Tl	203	65.0	35.5	**0.0028**	0.003	94.5	ug/L	31	Standard
	Tl	205	33.3	45.8	**0.0022**	0.002	93.6	ug/L	20	Standard
	Pb	206	995.4	6.0	**0.1186**	0.010	8.6	ug/L	286	Standard
	Pb	207	825.7	1.1	**0.1096**	0.003	2.3	ug/L	257	Standard
	Pb	208	3331.5	0.8	**0.1144**	0.002	1.9	ug/L	956	Standard
	U	238	237.0	6.0	**0.0298**	0.002	6.6	ug/L	17	Standard
>	**Bi**	209	562917.3	0.4				ug/L	557080	Standard

Sample ID: L1606114203
Report Date/Time: Tuesday, June 28, 2016 16:46:33
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	1456.7	5.2	**1.8709**	0.119	6.4	mg/L	32	Standard
K	39	195.0	2.6	**0.5136**	0.022	4.3	mg/L	30	Standard
Ca	43	66.7	37.7	**4.7333**	3.876	81.9	mg/L	48	Standard
Fe	54	1995.7	3.8	**1.6587**	0.089	5.4	mg/L	159	Standard
Fe	57	716.7	4.0	**1.7796**	0.068	3.8	mg/L	195	Standard
> **Sc-1**	45	39785.5	1.2				mg/L	39543	Standard
Cl	35	34494.6	2.9				ug/L	27642	Standard
Kr	83	1.3	173.2				ug/L	2	Standard
Br	81	3500.4	9.9				ug/L	1647	Standard
P	31	20160.1	3.5				ug/L	15231	Standard
S	34	2286.8	4.4				ug/L	1845	Standard
Sr	88	110.0	28.4				ug/L	70	Standard
C	12	216.7	25.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	341.7	16.9				mg/L	12	Standard
Ho-1	165	235.0	14.0				mg/L	12	Standard
Er	166	243.3	11.9				mg/L	20	Standard
I	127	25506.6	11.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		100.285	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.945	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606114203
Report Date/Time: Tuesday, June 28, 2016 16:46:33
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.342
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.048
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606114205
Sample Date/Time: Tuesday, June 28, 2016 16:47:28
Number of Replicates: 3
Autosampler Position: 245
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82787.3	1.4				ug/L	81760	Standard
	Be	9	48.3	23.9	**0.0383**	0.011	29.1	ug/L	17	Standard
	Al	27	819185.4	1.5	**5.2350**	0.125	2.4	ug/L	1410	Standard
	Sc	45	40940.2	1.5				ug/L	39543	Standard
	Ti	47	480.0	5.9	**1.6939**	0.112	6.6	ug/L	39	Standard
	V	51	6800.6	1.0	**0.7356**	0.007	0.9	ug/L	1415	Standard
	Cr	52	13980.4	1.6	**0.8771**	0.035	4.0	ug/L	6946	Standard
	Cr	53	1753.4	3.0	**1.5823**	0.062	3.9	ug/L	328	Standard
	Mn	55	596277.6	1.3	**88.4306**	1.209	1.4	ug/L	823	Standard
	Co	59	3687.4	0.3	**0.4741**	0.004	0.7	ug/L	252	Standard
	Ni	60	3717.1	1.3	**1.7958**	0.025	1.4	ug/L	192	Standard
	Cu	65	1213.0	2.2	**0.5240**	0.015	2.9	ug/L	141	Standard
	Zn	66	2977.6	2.2	**1.7886**	0.052	2.9	ug/L	491	Standard
>	**Ge**	72	534849.8	0.4				ug/L	515087	Standard
	As	75	377.9	7.0	**0.3640**	0.023	6.2	ug/L	20	Standard
	Se	82	30.4	15.6	**0.1347**	0.040	29.6	ug/L	15	Standard
	Se-1	77	66.7	7.4	**0.1333**	0.063	47.3	ug/L	48	Standard
>	**Ga**	71	223.3	16.2				mg/L	90	Standard
	Rb	85	9057.7	1.9				ug/L	23	Standard
	Y	89	448330.1	1.0				ug/L	448510	Standard
>	**Rh**	103	13.3	21.7				ug/L	15	Standard
	Mo	98	830.0	4.5	**0.2872**	0.016	5.5	ug/L	11	Standard
	Ag	107	122.7	18.8	**0.0020**	0.004	175.3	ug/L	134	Standard
	Cd	111	156.2	7.4	**0.0700**	0.006	8.4	mg/L	7	Standard
	Cd	114	373.1	12.2	**0.0667**	0.009	13.8	ug/L	16	Standard
>	**In**	115	612810.6	0.7				ug/L	611117	Standard
	Sn	118	755.0	9.6	**0.0254**	0.012	48.6	ug/L	518	Standard
	Sb	123	285.8	9.7	**0.0476**	0.006	12.8	ug/L	127	Standard
	Ba	135	48731.6	0.5	**23.5144**	0.057	0.2	ug/L	45	Standard
	Ce	140	12518.5	3.0				ug/L	298	Standard
>	**Tb**	159	979236.7	1.3				ug/L	977942	Standard
	Ho	165	273.3	12.2				ug/L	12	Standard
	Tl	203	95.3	25.1	**0.0062**	0.003	47.7	ug/L	31	Standard
	Tl	205	61.7	4.7	**0.0059**	0.000	5.9	ug/L	20	Standard
	Pb	206	1176.7	3.8	**0.1481**	0.011	7.4	ug/L	286	Standard
	Pb	207	985.7	2.0	**0.1385**	0.005	4.0	ug/L	257	Standard
	Pb	208	4033.5	3.6	**0.1475**	0.011	7.5	ug/L	956	Standard
	U	238	279.0	8.7	**0.0349**	0.004	10.9	ug/L	17	Standard
>	**Bi**	209	573358.0	1.7				ug/L	557080	Standard

Sample ID: L1606114205
Report Date/Time: Tuesday, June 28, 2016 16:49:45
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.1692**	17.605	173.1	mg/L	0	Standard
Mg	24	1471.7	7.4	**1.8343**	0.119	6.5	mg/L	32	Standard
K	39	181.7	21.4	**0.4516**	0.129	28.5	mg/L	30	Standard
Ca	43	60.0	36.3	**3.4344**	3.118	90.8	mg/L	48	Standard
Fe	54	2300.2	8.7	**1.8724**	0.148	7.9	mg/L	159	Standard
Fe	57	716.7	8.3	**1.7108**	0.158	9.3	mg/L	195	Standard
> **Sc-1**	45	40940.2	1.5				mg/L	39543	Standard
Cl	35	34657.6	1.8				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	3160.3	4.1				ug/L	1647	Standard
P	31	20585.6	2.1				ug/L	15231	Standard
S	34	2273.5	1.5				ug/L	1845	Standard
Sr	88	120.0	11.0				ug/L	70	Standard
C	12	246.7	23.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	334.1	26.2				mg/L	12	Standard
Ho-1	165	273.3	12.2				mg/L	12	Standard
Er	166	263.3	24.4				mg/L	20	Standard
I	127	21154.8	6.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.256	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.837	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606114205
Report Date/Time: Tuesday, June 28, 2016 16:49:45
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.277
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.922
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115102
Sample Date/Time: Tuesday, June 28, 2016 16:50:39
Number of Replicates: 3
Autosampler Position: 246
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83962.1	2.2				ug/L	81760	Standard
	Be	9	23.3	32.7	**0.0132**	0.008	57.7	ug/L	17	Standard
	Al	27	14475219.3	2.2	**91.4576**	2.716	3.0	ug/L	1410	Standard
	Sc	45	40821.6	0.4				ug/L	39543	Standard
	Ti	47	413.3	6.2	**1.5180**	0.138	9.1	ug/L	39	Standard
	V	51	16629.8	2.6	**2.2151**	0.113	5.1	ug/L	1415	Standard
	Cr	52	10947.0	2.1	**0.5345**	0.067	12.6	ug/L	6946	Standard
	Cr	53	3965.5	4.7	**4.2265**	0.126	3.0	ug/L	328	Standard
	Mn	55	185468.7	3.9	**28.8128**	1.697	5.9	ug/L	823	Standard
	Co	59	997.7	2.0	**0.1124**	0.006	4.9	ug/L	252	Standard
	Ni	60	1478.4	4.6	**0.6898**	0.053	7.7	ug/L	192	Standard
	Cu	65	1958.8	1.9	**0.9525**	0.040	4.2	ug/L	141	Standard
	Zn	66	8106.5	2.8	**6.6572**	0.319	4.8	ug/L	491	Standard
>	**Ge**	72	509236.4	2.2				ug/L	515087	Standard
	As	75	962.9	2.3	**0.9070**	0.035	3.9	ug/L	20	Standard
	Se	82	228.1	4.7	**1.8861**	0.120	6.4	ug/L	15	Standard
	Se-1	77	223.7	5.0	**2.2872**	0.212	9.3	ug/L	48	Standard
>	**Ga**	71	265.0	37.9				mg/L	90	Standard
	Rb	85	21039.6	2.1				ug/L	23	Standard
	Y	89	444144.2	0.6				ug/L	448510	Standard
>	**Rh**	103	50.0	20.0				ug/L	15	Standard
	Mo	98	3394.4	4.1	**1.2339**	0.051	4.2	ug/L	11	Standard
	Ag	107	107.3	13.1	**-0.0003**	0.002	751.8	ug/L	134	Standard
	Cd	111	62.6	4.9	**0.0263**	0.001	5.7	mg/L	7	Standard
	Cd	114	195.9	21.1	**0.0327**	0.008	24.5	ug/L	16	Standard
>	**In**	115	612517.0	0.3				ug/L	611117	Standard
	Sn	118	836.7	5.7	**0.0392**	0.008	20.1	ug/L	518	Standard
	Sb	123	547.1	4.7	**0.1028**	0.005	5.0	ug/L	127	Standard
	Ba	135	71666.8	2.3	**34.6081**	0.699	2.0	ug/L	45	Standard
	Ce	140	3362.0	2.6				ug/L	298	Standard
>	**Tb**	159	988372.6	0.5				ug/L	977942	Standard
	Ho	165	63.3	75.9				ug/L	12	Standard
	Tl	203	105.7	10.4	**0.0078**	0.001	16.0	ug/L	31	Standard
	Tl	205	108.3	39.8	**0.0125**	0.006	45.9	ug/L	20	Standard
	Pb	206	1800.1	2.5	**0.2697**	0.008	3.1	ug/L	286	Standard
	Pb	207	1518.4	2.6	**0.2536**	0.007	2.9	ug/L	257	Standard
	Pb	208	6212.5	1.8	**0.2692**	0.011	4.0	ug/L	956	Standard
	U	238	10075.0	2.3	**1.4043**	0.047	3.4	ug/L	17	Standard
>	**Bi**	209	556782.3	1.6				ug/L	557080	Standard

Sample ID: L1606115102
Report Date/Time: Tuesday, June 28, 2016 16:52:56
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**20.7586**	17.973	86.6	mg/L	0	Standard
Mg	24	36599.2	3.6	**47.1388**	1.781	3.8	mg/L	32	Standard
K	39	311.7	20.3	**0.8782**	0.204	23.2	mg/L	30	Standard
Ca	43	200.0	10.0	**24.0730**	2.845	11.8	mg/L	48	Standard
Fe	54	272.4	27.9	**0.1018**	0.067	65.8	mg/L	159	Standard
Fe	57	400.0	17.2	**0.6830**	0.221	32.3	mg/L	195	Standard
Sc-1	45	40821.6	0.4				mg/L	39543	Standard
Cl	35	37764.4	4.8				ug/L	27642	Standard
Kr	83	2.3	49.5				ug/L	2	Standard
Br	81	41179.3	5.9				ug/L	1647	Standard
P	31	20392.0	1.6				ug/L	15231	Standard
S	34	2350.2	4.9				ug/L	1845	Standard
Sr	88	246.7	3.1				ug/L	70	Standard
C	12	240.0	16.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	93.2	27.3				mg/L	12	Standard
Ho-1	165	63.3	75.9				mg/L	12	Standard
Er	166	73.3	7.9				mg/L	20	Standard
I	127	267945.4	8.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		102.693	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		98.864	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606115102
Report Date/Time: Tuesday, June 28, 2016 16:52:56
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.229
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.946
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115105
Sample Date/Time: Tuesday, June 28, 2016 16:53:51
Number of Replicates: 3
Autosampler Position: 247
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81681.4	3.1				ug/L	81760	Standard
	Be	9	16.7	45.8	**0.0069**	0.007	104.2	ug/L	17	Standard
	Al	27	18070414.8	1.1	**117.3803**	2.444	2.1	ug/L	1410	Standard
	Sc	45	40721.3	2.6				ug/L	39543	Standard
	Ti	47	244.7	8.0	**0.8457**	0.069	8.1	ug/L	39	Standard
	V	51	17618.4	3.0	**2.3985**	0.046	1.9	ug/L	1415	Standard
	Cr	52	10888.2	1.3	**0.5506**	0.002	0.4	ug/L	6946	Standard
	Cr	53	4290.6	7.0	**4.6855**	0.402	8.6	ug/L	328	Standard
	Mn	55	234137.7	1.0	**36.9743**	0.056	0.2	ug/L	823	Standard
	Co	59	907.0	1.1	**0.1013**	0.001	0.9	ug/L	252	Standard
	Ni	60	1595.4	2.1	**0.7658**	0.017	2.3	ug/L	192	Standard
	Cu	65	1678.8	2.6	**0.8172**	0.017	2.1	ug/L	141	Standard
	Zn	66	5525.3	1.2	**4.3540**	0.048	1.1	ug/L	491	Standard
>	**Ge**	72	501124.8	1.2				ug/L	515087	Standard
	As	75	1072.7	4.1	**1.0212**	0.046	4.5	ug/L	20	Standard
	Se	82	264.0	5.9	**2.2378**	0.139	6.2	ug/L	15	Standard
	Se-1	77	284.0	16.7	**3.1586**	0.667	21.1	ug/L	48	Standard
>	**Ga**	71	116.7	13.8				mg/L	90	Standard
	Rb	85	15923.4	4.0				ug/L	23	Standard
	Y	89	444308.5	1.8				ug/L	448510	Standard
>	**Rh**	103	68.3	40.3				ug/L	15	Standard
	Mo	98	3823.6	1.7	**1.4182**	0.006	0.4	ug/L	11	Standard
	Ag	107	76.3	4.0	**-0.0047**	0.001	10.8	ug/L	134	Standard
	Cd	111	40.8	13.4	**0.0165**	0.003	15.5	mg/L	7	Standard
	Cd	114	111.3	16.3	**0.0169**	0.004	22.4	ug/L	16	Standard
>	**In**	115	601437.2	1.6				ug/L	611117	Standard
	Sn	118	1256.7	47.3	**0.1130**	0.098	86.7	ug/L	518	Standard
	Sb	123	502.0	6.8	**0.0953**	0.009	9.3	ug/L	127	Standard
	Ba	135	85456.0	1.5	**42.0358**	0.440	1.0	ug/L	45	Standard
	Ce	140	1221.7	20.4				ug/L	298	Standard
>	**Tb**	159	951914.7	0.8				ug/L	977942	Standard
	Ho	165	38.3	39.8				ug/L	12	Standard
	Tl	203	109.0	4.0	**0.0087**	0.001	7.8	ug/L	31	Standard
	Tl	205	100.0	13.2	**0.0120**	0.002	15.9	ug/L	20	Standard
	Pb	206	942.7	3.4	**0.1187**	0.004	3.7	ug/L	286	Standard
	Pb	207	751.4	3.5	**0.1031**	0.007	7.3	ug/L	257	Standard
	Pb	208	3320.1	1.9	**0.1235**	0.005	4.4	ug/L	956	Standard
	U	238	12716.0	1.7	**1.8521**	0.048	2.6	ug/L	17	Standard
>	**Bi**	209	533061.4	1.1				ug/L	557080	Standard

Sample ID: L1606115105
Report Date/Time: Tuesday, June 28, 2016 16:56:08
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.4978**	18.174	173.1	mg/L	0	Standard
Mg	24	45127.2	3.7	**58.2718**	1.297	2.2	mg/L	32	Standard
K	39	268.3	17.6	**0.7391**	0.156	21.1	mg/L	30	Standard
Ca	43	230.0	5.8	**28.6229**	2.753	9.6	mg/L	48	Standard
Fe	54	227.2	16.5	**0.0631**	0.037	58.2	mg/L	159	Standard
Fe	57	321.7	8.6	**0.4294**	0.079	18.3	mg/L	195	Standard
> **Sc-1**	45	40721.3	2.6				mg/L	39543	Standard
Cl	35	40535.5	2.1				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	42472.8	3.2				ug/L	1647	Standard
P	31	21144.7	2.2				ug/L	15231	Standard
S	34	2520.2	4.3				ug/L	1845	Standard
Sr	88	268.3	23.3				ug/L	70	Standard
C	12	270.0	30.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	48.6	21.1				mg/L	12	Standard
Ho-1	165	38.3	39.8				mg/L	12	Standard
Er	166	30.0	33.3				mg/L	20	Standard
I	127	357293.3	4.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		99.904	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.289	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115105
Report Date/Time: Tuesday, June 28, 2016 16:56:08
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.416
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.688
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606115105
Report Date/Time: Tuesday, June 28, 2016 16:56:08
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115401 WG573751-03
Sample Date/Time: Tuesday, June 28, 2016 16:57:02
Number of Replicates: 3
Autosampler Position: 248
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100430.3	2.8				ug/L	81760	Standard
	Be	9	28.3	56.7	**0.0134**	0.013	96.3	ug/L	17	Standard
	Al	27	81335563.7	1.7	**429.6767**	8.046	1.9	ug/L	1410	Standard
	Sc	45	36054.5	1.3				ug/L	39543	Standard
	Ti	47	276.7	4.4	**1.0861**	0.041	3.8	ug/L	39	Standard
	V	51	56034.7	2.7	**8.8625**	0.243	2.7	ug/L	1415	Standard
	Cr	52	12377.7	0.7	**0.9433**	0.019	2.0	ug/L	6946	Standard
	Cr	53	23940.6	3.3	**30.3443**	1.215	4.0	ug/L	328	Standard
	Mn	55	741944.6	0.9	**128.6152**	2.806	2.2	ug/L	823	Standard
	Co	59	3502.1	2.3	**0.5295**	0.021	4.0	ug/L	252	Standard
	Ni	60	10135.4	1.6	**5.9492**	0.102	1.7	ug/L	192	Standard
	Cu	65	4993.5	1.7	**2.8661**	0.085	3.0	ug/L	141	Standard
	Zn	66	5821.5	1.7	**5.1486**	0.184	3.6	ug/L	491	Standard
>	**Ge**	72	457910.9	1.5				ug/L	515087	Standard
	As	75	1957.2	6.1	**1.9997**	0.145	7.2	ug/L	20	Standard
	Se	82	3071.1	3.5	**29.9014**	1.486	5.0	ug/L	15	Standard
	Se-1	77	4244.3	1.8	**62.6742**	2.070	3.3	ug/L	48	Standard
>	**Ga**	71	138.3	10.4				mg/L	90	Standard
	Rb	85	9993.3	0.6				ug/L	23	Standard
	Y	89	397349.5	0.4				ug/L	448510	Standard
>	**Rh**	103	1400.1	6.9				ug/L	15	Standard
	Mo	98	28731.5	0.6	**12.2349**	0.110	0.9	ug/L	11	Standard
	Ag	107	395.7	2.6	**0.0517**	0.002	4.3	ug/L	134	Standard
	Cd	111	92.0	15.6	**0.0467**	0.008	17.2	mg/L	7	Standard
	Cd	114	319.4	5.6	**0.0659**	0.004	5.6	ug/L	16	Standard
>	**In**	115	530125.1	0.7				ug/L	611117	Standard
	Sn	118	1140.0	5.3	**0.1203**	0.011	8.8	ug/L	518	Standard
	Sb	123	532.5	6.0	**0.1172**	0.007	5.9	ug/L	127	Standard
	Ba	135	16736.6	1.8	**9.3206**	0.182	2.0	ug/L	45	Standard
	Ce	140	1620.1	9.0				ug/L	298	Standard
>	**Tb**	159	890828.2	2.0				ug/L	977942	Standard
	Ho	165	38.3	27.2				ug/L	12	Standard
	Tl	203	267.7	7.7	**0.0327**	0.003	10.1	ug/L	31	Standard
	Tl	205	231.7	6.6	**0.0350**	0.002	6.9	ug/L	20	Standard
	Pb	206	883.7	4.8	**0.1290**	0.007	5.4	ug/L	286	Standard
	Pb	207	738.7	4.5	**0.1206**	0.010	8.0	ug/L	257	Standard
	Pb	208	2811.4	0.6	**0.1152**	0.004	3.5	ug/L	956	Standard
	U	238	63722.1	1.7	**10.4767**	0.095	0.9	ug/L	17	Standard
>	**Bi**	209	472777.8	1.7				ug/L	557080	Standard

Sample ID: L1606115401 WG573751-03
Report Date/Time: Tuesday, June 28, 2016 16:59:19
Page 1

Approved: June 29, 2016

Na	23	15.0	33.3	**105.9031**	36.283	34.3	mg/L	0	Standard
Mg	24	356514.1	0.8	**520.5310**	10.800	2.1	mg/L	32	Standard
K	39	430.0	2.3	**1.4525**	0.056	3.9	mg/L	30	Standard
Ca	43	1840.1	5.6	**301.0738**	13.488	4.5	mg/L	48	Standard
Fe	54	188.6	12.1	**0.0500**	0.021	41.8	mg/L	159	Standard
Fe	57	1585.1	2.5	**5.2457**	0.200	3.8	mg/L	195	Standard
> **Sc-1**	45	36054.5	1.3				mg/L	39543	Standard
Cl	35	90598.1	1.3				ug/L	27642	Standard
Kr	83	3.0	66.7				ug/L	2	Standard
Br	81	42312.4	3.9				ug/L	1647	Standard
P	31	19347.4	2.8				ug/L	15231	Standard
S	34	2421.9	2.4				ug/L	1845	Standard
Sr	88	6231.3	0.9				ug/L	70	Standard
C	12	243.3	12.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	37.9	25.6				mg/L	12	Standard
Ho-1	165	38.3	27.2				mg/L	12	Standard
Er	166	43.3	96.1				mg/L	20	Standard
I	127	118064.4	6.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		122.835	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		88.900	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115401 WG573751-03
Report Date/Time: Tuesday, June 28, 2016 16:59:19
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	86.747
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	84.867
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606115401 WG573751-03
Report Date/Time: Tuesday, June 28, 2016 16:59:19
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115402S WG573751-09
Sample Date/Time: Tuesday, June 28, 2016 17:00:14
Number of Replicates: 3
Autosampler Position: 249
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	111794.4	2.2				ug/L	81760	Standard
	Be	9	61074.6	3.8	**44.9573**	0.853	1.9	ug/L	17	Standard
	Al	27	84589452.8	4.6	**401.1886**	9.792	2.4	ug/L	1410	Standard
	Sc	45	39148.9	4.1				ug/L	39543	Standard
	Ti	47	272.0	24.8	**0.9934**	0.318	32.0	ug/L	39	Standard
	V	51	406216.4	3.4	**61.4172**	0.609	1.0	ug/L	1415	Standard
	Cr	52	347340.1	3.9	**51.5532**	0.830	1.6	ug/L	6946	Standard
	Cr	53	63548.0	2.3	**76.0057**	0.198	0.3	ug/L	328	Standard
	Mn	55	986761.6	2.9	**160.3601**	1.162	0.7	ug/L	823	Standard
	Co	59	344414.5	2.6	**51.6648**	0.567	1.1	ug/L	252	Standard
	Ni	60	96461.7	3.6	**53.9005**	0.670	1.2	ug/L	192	Standard
	Cu	65	88941.7	2.7	**49.2712**	0.171	0.3	ug/L	141	Standard
	Zn	66	55366.0	3.1	**52.4384**	0.443	0.8	ug/L	491	Standard
>	**Ge**	72	488431.6	2.5				ug/L	515087	Standard
	As	75	59145.7	4.0	**55.5035**	0.877	1.6	ug/L	20	Standard
	Se	82	9257.9	4.7	**84.6512**	1.891	2.2	ug/L	15	Standard
	Se-1	77	8241.9	2.0	**114.6628**	0.599	0.5	ug/L	48	Standard
>	**Ga**	71	135.0	3.7				mg/L	90	Standard
	Rb	85	11232.5	8.2				ug/L	23	Standard
	Y	89	424780.5	4.2				ug/L	448510	Standard
>	**Rh**	103	1410.1	5.3				ug/L	15	Standard
	Mo	98	28554.4	4.0	**11.2353**	0.049	0.4	ug/L	11	Standard
	Ag	107	298500.0	4.4	**47.3531**	0.325	0.7	ug/L	134	Standard
	Cd	111	96838.7	4.3	**48.2572**	0.264	0.5	mg/L	7	Standard
	Cd	114	236461.6	3.9	**48.4818**	0.307	0.6	ug/L	16	Standard
>	**In**	115	573601.2	3.7				ug/L	611117	Standard
	Sn	118	926.7	16.3	**0.0649**	0.025	39.2	ug/L	518	Standard
	Sb	123	232072.7	4.2	**52.3852**	0.509	1.0	ug/L	127	Standard
	Ba	135	113861.5	4.7	**58.7202**	0.562	1.0	ug/L	45	Standard
	Ce	140	1120.0	4.0				ug/L	298	Standard
>	**Tb**	159	960238.4	4.5				ug/L	977942	Standard
	Ho	165	33.3	75.5				ug/L	12	Standard
	Tl	203	395828.3	3.1	**51.4711**	0.762	1.5	ug/L	31	Standard
	Tl	205	346560.1	3.5	**51.8878**	0.159	0.3	ug/L	20	Standard
	Pb	206	263890.3	2.7	**53.4438**	0.979	1.8	ug/L	286	Standard
	Pb	207	230261.5	2.5	**51.7030**	0.928	1.8	ug/L	257	Standard
	Pb	208	887552.6	2.9	**51.5030**	0.554	1.1	ug/L	956	Standard
	U	238	419873.0	2.9	**64.2072**	0.642	1.0	ug/L	17	Standard
>	**Bi**	209	508536.8	3.8				ug/L	557080	Standard

Sample ID: L1606115402S WG573751-09
Report Date/Time: Tuesday, June 28, 2016 17:02:31
Page 1

Approved: June 29, 2016

Na	23	21.7	35.3	**139.6053**	45.304	32.5	mg/L	0	Standard
Mg	24	357824.7	2.8	**481.2852**	10.577	2.2	mg/L	32	Standard
K	39	491.7	6.5	**1.5399**	0.155	10.1	mg/L	30	Standard
Ca	43	1775.1	1.7	**267.3999**	15.688	5.9	mg/L	48	Standard
Fe	54	171.9	15.3	**0.0199**	0.020	102.8	mg/L	159	Standard
Fe	57	1550.1	0.6	**4.6666**	0.208	4.5	mg/L	195	Standard
> **Sc-1**	45	39148.9	4.1				mg/L	39543	Standard
Cl	35	91435.3	3.5				ug/L	27642	Standard
Kr	83	2.7	21.7				ug/L	2	Standard
Br	81	45418.1	6.1				ug/L	1647	Standard
P	31	20330.3	5.5				ug/L	15231	Standard
S	34	2288.5	3.3				ug/L	1845	Standard
Sr	88	6416.4	6.4				ug/L	70	Standard
C	12	206.7	37.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	21.9	50.9				mg/L	12	Standard
Ho-1	165	33.3	75.5				mg/L	12	Standard
Er	166	30.0	57.7				mg/L	20	Standard
I	127	49499.1	2.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		136.735	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.825	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115402S WG573751-09
Report Date/Time: Tuesday, June 28, 2016 17:02:31
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	93.861
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	91.286
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606115402S WG573751-09
Report Date/Time: Tuesday, June 28, 2016 17:02:31
Page 3

Approved: June 29, 2016

Se-1 77 Upper, S, EEE	Se-1	77	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115403SD WG573751-10
Sample Date/Time: Tuesday, June 28, 2016 17:03:25
Number of Replicates: 3
Autosampler Position: 250
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	118994.0	2.6				ug/L	81760	Standard
	Be	9	64654.3	0.9	**44.7357**	0.820	1.8	ug/L	17	Standard
	Al	27	87741403.7	3.4	**391.0655**	6.582	1.7	ug/L	1410	Standard
	Sc	45	41172.5	1.6				ug/L	39543	Standard
	Ti	47	259.3	3.7	**0.9238**	0.057	6.2	ug/L	39	Standard
	V	51	425385.5	3.2	**63.6566**	1.295	2.0	ug/L	1415	Standard
	Cr	52	362069.3	1.9	**53.2267**	0.604	1.1	ug/L	6946	Standard
	Cr	53	64441.8	2.0	**76.2717**	0.872	1.1	ug/L	328	Standard
	Mn	55	1073606.3	3.1	**172.6721**	4.096	2.4	ug/L	823	Standard
	Co	59	369825.3	3.2	**54.8942**	1.128	2.1	ug/L	252	Standard
	Ni	60	100548.2	1.5	**55.6131**	0.149	0.3	ug/L	192	Standard
	Cu	65	92359.0	1.3	**50.6385**	0.319	0.6	ug/L	141	Standard
	Zn	66	57865.3	2.6	**54.2651**	0.864	1.6	ug/L	491	Standard
>	**Ge**	72	493572.0	1.8				ug/L	515087	Standard
	As	75	61885.0	3.3	**57.4777**	1.381	2.4	ug/L	20	Standard
	Se	82	9664.4	3.7	**87.4717**	2.250	2.6	ug/L	15	Standard
	Se-1	77	8366.3	3.5	**115.1653**	2.897	2.5	ug/L	48	Standard
>	**Ga**	71	118.3	21.7				mg/L	90	Standard
	Rb	85	11170.8	6.4				ug/L	23	Standard
	Y	89	447220.0	3.1				ug/L	448510	Standard
>	**Rh**	103	1640.1	8.8				ug/L	15	Standard
	Mo	98	29726.0	4.1	**11.0857**	0.160	1.4	ug/L	11	Standard
	Ag	107	312572.9	2.2	**47.0220**	0.762	1.6	ug/L	134	Standard
	Cd	111	102356.1	3.3	**48.3584**	0.794	1.6	mg/L	7	Standard
	Cd	114	246212.9	4.6	**47.8408**	0.884	1.8	ug/L	16	Standard
>	**In**	115	605206.5	3.7				ug/L	611117	Standard
	Sn	118	1233.4	10.6	**0.1085**	0.018	16.3	ug/L	518	Standard
	Sb	123	240651.4	1.9	**51.5127**	0.935	1.8	ug/L	127	Standard
	Ba	135	117995.0	3.4	**57.6913**	0.169	0.3	ug/L	45	Standard
	Ce	140	1415.1	2.3				ug/L	298	Standard
>	**Tb**	159	994298.0	3.0				ug/L	977942	Standard
	Ho	165	46.7	12.4				ug/L	12	Standard
	Tl	203	412661.0	1.3	**51.8933**	0.197	0.4	ug/L	31	Standard
	Tl	205	366877.2	3.6	**53.1179**	1.107	2.1	ug/L	20	Standard
	Pb	206	275320.1	2.3	**53.9118**	0.413	0.8	ug/L	286	Standard
	Pb	207	238830.1	2.6	**51.8469**	0.552	1.1	ug/L	257	Standard
	Pb	208	936516.1	2.9	**52.5472**	0.756	1.4	ug/L	956	Standard
	U	238	442147.9	3.9	**65.3698**	1.598	2.4	ug/L	17	Standard
>	**Bi**	209	525736.2	1.5				ug/L	557080	Standard

Sample ID: L1606115403SD WG573751-10
Report Date/Time: Tuesday, June 28, 2016 17:05:42
Page 1

Approved: June 29, 2016

Na	23	31.7	48.2	**194.7966**	93.559	48.0	mg/L	0	Standard
Mg	24	381865.2	2.0	**488.2175**	12.226	2.5	mg/L	32	Standard
K	39	475.0	11.9	**1.3985**	0.159	11.4	mg/L	30	Standard
Ca	43	1828.4	7.5	**261.2707**	17.620	6.7	mg/L	48	Standard
Fe	54	210.4	5.2	**0.0459**	0.012	26.9	mg/L	159	Standard
Fe	57	1693.4	5.0	**4.8640**	0.189	3.9	mg/L	195	Standard
> **Sc-1**	45	41172.5	1.6				mg/L	39543	Standard
Cl	35	96785.1	3.5				ug/L	27642	Standard
Kr	83	2.7	78.1				ug/L	2	Standard
Br	81	48665.0	7.0				ug/L	1647	Standard
P	31	21074.7	3.4				ug/L	15231	Standard
S	34	2411.9	0.6				ug/L	1845	Standard
Sr	88	7001.6	9.4				ug/L	70	Standard
C	12	253.3	28.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	54.1	45.2				mg/L	12	Standard
Ho-1	165	46.7	12.4				mg/L	12	Standard
Er	166	53.3	28.6				mg/L	20	Standard
I	127	96813.1	12.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		145.540	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.823	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115403SD WG573751-10
Report Date/Time: Tuesday, June 28, 2016 17:05:42
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.033
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.373
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606115403SD WG573751-10
Report Date/Time: Tuesday, June 28, 2016 17:05:42
Page 3

Approved: June 29, 2016

Se-1 77 Upper, S, EEE	Se-1	77	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115404
Sample Date/Time: Tuesday, June 28, 2016 17:06:37
Number of Replicates: 3
Autosampler Position: 251
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100804.3	1.4				ug/L	81760	Standard
	Be	9	43.3	6.7	**0.0257**	0.002	9.1	ug/L	17	Standard
	Al	27	95134.9	5.2	**0.4862**	0.026	5.4	ug/L	1410	Standard
	Sc	45	38767.9	2.5				ug/L	39543	Standard
	Ti	47	56.0	17.6	**0.0568**	0.042	74.0	ug/L	39	Standard
	V	51	2653.2	7.2	**0.1688**	0.023	13.6	ug/L	1415	Standard
	Cr	52	16846.0	0.3	**1.3242**	0.040	3.0	ug/L	6946	Standard
	Cr	53	2448.5	10.3	**2.4055**	0.318	13.2	ug/L	328	Standard
	Mn	55	6416.4	0.3	**0.8254**	0.012	1.5	ug/L	823	Standard
	Co	59	385.0	5.1	**0.0226**	0.003	13.2	ug/L	252	Standard
	Ni	60	1595.4	9.2	**0.7282**	0.071	9.8	ug/L	192	Standard
	Cu	65	1383.7	5.1	**0.6260**	0.046	7.3	ug/L	141	Standard
	Zn	66	3329.0	1.8	**2.1614**	0.092	4.3	ug/L	491	Standard
>	**Ge**	72	523540.2	1.3				ug/L	515087	Standard
	As	75	21.6	288.4	**0.0594**	0.054	91.4	ug/L	20	Standard
	Se	82	48.4	30.2	**0.2956**	0.131	44.3	ug/L	15	Standard
	Se-1	77	118.7	6.2	**0.8319**	0.117	14.0	ug/L	48	Standard
>	**Ga**	71	35.0	51.5				mg/L	90	Standard
	Rb	85	285.0	10.7				ug/L	23	Standard
	Y	89	449294.4	0.5				ug/L	448510	Standard
>	**Rh**	103	35.0	14.3				ug/L	15	Standard
	Mo	98	461.6	2.9	**0.1485**	0.005	3.0	ug/L	11	Standard
	Ag	107	143.3	7.2	**0.0047**	0.002	32.8	ug/L	134	Standard
	Cd	111	43.9	7.3	**0.0172**	0.002	8.8	mg/L	7	Standard
	Cd	114	151.5	5.9	**0.0237**	0.002	7.4	ug/L	16	Standard
>	**In**	115	622732.4	0.3				ug/L	611117	Standard
	Sn	118	3920.5	2.1	**0.5488**	0.015	2.8	ug/L	518	Standard
	Sb	123	270.3	27.2	**0.0434**	0.015	35.5	ug/L	127	Standard
	Ba	135	1327.1	1.7	**0.6057**	0.011	1.9	ug/L	45	Standard
	Ce	140	168.3	16.4				ug/L	298	Standard
>	**Tb**	159	980536.6	0.9				ug/L	977942	Standard
	Ho	165	21.7	13.3				ug/L	12	Standard
	Tl	203	266.3	8.1	**0.0259**	0.003	10.5	ug/L	31	Standard
	Tl	205	201.7	5.7	**0.0244**	0.001	5.0	ug/L	20	Standard
	Pb	206	459.0	5.5	**0.0188**	0.005	29.2	ug/L	286	Standard
	Pb	207	389.7	6.3	**0.0195**	0.006	29.5	ug/L	257	Standard
	Pb	208	1625.7	6.2	**0.0231**	0.006	26.8	ug/L	956	Standard
	U	238	121.7	7.4	**0.0135**	0.001	9.9	ug/L	17	Standard
>	**Bi**	209	574701.5	1.2				ug/L	557080	Standard

Sample ID: L1606115404
Report Date/Time: Tuesday, June 28, 2016 17:08:54
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**32.7163**	32.227	98.5	mg/L	0	Standard
Mg	24	330.0	22.9	**0.3898**	0.094	24.2	mg/L	32	Standard
K	39	33.3	43.3	**-0.0268**	0.048	179.7	mg/L	30	Standard
Ca	43	33.3	43.3	**-0.2011**	2.146	1067.4	mg/L	48	Standard
Fe	54	141.8	17.1	**-0.0059**	0.025	416.6	mg/L	159	Standard
Fe	57	216.7	23.3	**0.1207**	0.166	137.9	mg/L	195	Standard
Sc-1	45	38767.9	2.5				mg/L	39543	Standard
Cl	35	43132.0	0.5				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	3107.0	8.7				ug/L	1647	Standard
P	31	17094.6	3.8				ug/L	15231	Standard
S	34	2171.8	4.7				ug/L	1845	Standard
Sr	88	71.7	31.5				ug/L	70	Standard
C	12	180.0	44.1				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	6.3	86.6				mg/L	12	Standard
Ho-1	165	21.7	13.3				mg/L	12	Standard
Er	166	6.7	86.6				mg/L	20	Standard
I	127	6683.3	38.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		123.293	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		101.641	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606115404
Report Date/Time: Tuesday, June 28, 2016 17:08:54
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.901
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.163
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 17:09:50
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98106.9	3.9				ug/L	81760	Standard
	Be	9	60781.9	5.8	**50.9762**	1.224	2.4	ug/L	17	Standard
	Al	27	8271215.1	2.8	**44.7180**	0.568	1.3	ug/L	1410	Standard
	Sc	45	39052.0	5.6				ug/L	39543	Standard
	Ti	47	24646.4	4.4	**101.4481**	2.619	2.6	ug/L	39	Standard
	V	51	347970.7	1.4	**51.1411**	0.534	1.0	ug/L	1415	Standard
	Cr	52	354198.6	3.4	**51.1178**	0.779	1.5	ug/L	6946	Standard
	Cr	53	44595.6	4.3	**51.7475**	1.396	2.7	ug/L	328	Standard
	Mn	55	315347.9	1.9	**49.7410**	0.694	1.4	ug/L	823	Standard
	Co	59	365725.5	2.3	**53.3530**	0.828	1.6	ug/L	252	Standard
	Ni	60	93315.9	2.3	**50.7083**	0.301	0.6	ug/L	192	Standard
	Cu	65	94072.6	2.8	**50.6778**	0.428	0.8	ug/L	141	Standard
	Zn	66	55586.5	1.6	**51.1865**	0.432	0.8	ug/L	491	Standard
>	**Ge**	72	502280.1	2.3				ug/L	515087	Standard
	As	75	55868.4	1.2	**51.0087**	0.746	1.5	ug/L	20	Standard
	Se	82	5921.1	2.6	**52.6309**	1.371	2.6	ug/L	15	Standard
	Se-1	77	3936.2	0.2	**52.8821**	1.332	2.5	ug/L	48	Standard
>	**Ga**	71	43.3	6.7				mg/L	90	Standard
	Rb	85	853.4	12.0				ug/L	23	Standard
	Y	89	445806.9	0.5				ug/L	448510	Standard
>	**Rh**	103	45.0	61.9				ug/L	15	Standard
	Mo	98	273081.8	2.1	**98.9265**	1.774	1.8	ug/L	11	Standard
	Ag	107	339843.9	2.2	**49.5651**	0.762	1.5	ug/L	134	Standard
	Cd	111	106758.1	1.3	**48.9107**	0.367	0.7	mg/L	7	Standard
	Cd	114	258025.9	1.5	**48.6321**	0.403	0.8	ug/L	16	Standard
>	**In**	115	623979.7	0.7				ug/L	611117	Standard
	Sn	118	298451.0	1.1	**49.3312**	0.224	0.5	ug/L	518	Standard
	Sb	123	238781.4	2.0	**49.5506**	0.796	1.6	ug/L	127	Standard
	Ba	135	101752.9	1.4	**48.2444**	0.388	0.8	ug/L	45	Standard
	Ce	140	83.3	6.9				ug/L	298	Standard
>	**Tb**	159	999812.3	3.7				ug/L	977942	Standard
	Ho	165	18.3	56.8				ug/L	12	Standard
	Tl	203	438487.8	2.0	**50.5214**	0.372	0.7	ug/L	31	Standard
	Tl	205	396911.7	0.7	**52.6785**	1.200	2.3	ug/L	20	Standard
	Pb	206	278860.8	1.5	**50.0350**	0.542	1.1	ug/L	286	Standard
	Pb	207	252734.9	2.4	**50.2714**	0.461	0.9	ug/L	257	Standard
	Pb	208	1008405.1	1.0	**51.8566**	0.775	1.5	ug/L	956	Standard
	U	238	385967.4	0.9	**52.3060**	0.849	1.6	ug/L	17	Standard
>	**Bi**	209	573840.1	2.4				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:12:07
Page 1

Approved: June 29, 2016

Na	23	6.7	43.3	**42.7715**	15.957	37.3	mg/L	0	Standard
Mg	24	4330.6	4.7	**5.7850**	0.241	4.2	mg/L	32	Standard
K	39	1310.1	5.5	**4.3410**	0.087	2.0	mg/L	30	Standard
Ca	43	75.0	17.6	**6.2533**	2.500	40.0	mg/L	48	Standard
Fe	54	6082.2	3.3	**5.4477**	0.350	6.4	mg/L	159	Standard
Fe	57	1690.1	5.8	**5.1531**	0.140	2.7	mg/L	195	Standard
Sc-1	45	39052.0	5.6				mg/L	39543	Standard
Cl	35	36312.2	3.9				ug/L	27642	Standard
Kr	83	2.7	57.3				ug/L	2	Standard
Br	81	2130.1	3.3				ug/L	1647	Standard
P	31	18776.6	1.2				ug/L	15231	Standard
S	34	2388.5	6.0				ug/L	1845	Standard
Sr	88	65.0	13.3				ug/L	70	Standard
C	12	133.3	33.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	12.7	43.5				mg/L	12	Standard
Ho-1	165	18.3	56.8				mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	2708.6	18.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	101.952		
Al	27	89.436		
Sc	45			
Ti	47	101.448		
V	51	102.282		
Cr	52	102.236		
Cr	53			
Mn	55	99.482		
Co	59	106.706		
Ni	60	101.417		
Cu	65	101.356		
Zn	66	102.373		
Ge	72		97.514	
As	75	102.017		
Se	82	105.262		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:12:07
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	98.926		
Ag	107	99.130		
Cd	111	97.821		
Cd	114			
In	115		102.105	
Sn	118	98.662		
Sb	123	99.101		
Ba	135	96.489		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.043		
Tl	205			
Pb	206			
Pb	207			
Pb	208	103.713		
U	238	104.612		
Bi	209		103.008	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:12:07
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 17:13:01
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	92565.4	2.9				ug/L	81760	Standard
	Be	9	53.3	124.8	**0.0389**	0.062	158.6	ug/L	17	Standard
	Al	27	40870.6	146.9	**0.2266**	0.357	157.4	ug/L	1410	Standard
	Sc	45	37247.5	4.3				ug/L	39543	Standard
	Ti	47	28.7	28.0	**-0.0471**	0.033	70.8	ug/L	39	Standard
	V	51	1306.5	27.0	**-0.0143**	0.053	368.2	ug/L	1415	Standard
	Cr	52	6317.0	3.9	**-0.1289**	0.039	29.9	ug/L	6946	Standard
	Cr	53	723.4	11.4	**0.5015**	0.098	19.5	ug/L	328	Standard
	Mn	55	1084.7	54.0	**0.0228**	0.093	407.7	ug/L	823	Standard
	Co	59	338.3	37.8	**0.0179**	0.019	104.7	ug/L	252	Standard
	Ni	60	322.7	24.3	**0.0703**	0.043	61.2	ug/L	192	Standard
	Cu	65	237.7	21.2	**0.0371**	0.027	74.0	ug/L	141	Standard
	Zn	66	487.7	9.0	**-0.3725**	0.041	11.1	ug/L	491	Standard
>	Ge	72	503592.6	0.3				ug/L	515087	Standard
	As	75	-19.3	188.6	**0.0227**	0.033	146.0	ug/L	20	Standard
	Se	82	22.8	34.7	**0.0835**	0.070	83.8	ug/L	15	Standard
	Se-1	77	71.3	3.2	**0.2496**	0.032	12.8	ug/L	48	Standard
>	Ga	71	43.3	46.6				mg/L	90	Standard
	Rb	85	33.3	37.7				ug/L	23	Standard
	Y	89	434857.2	0.6				ug/L	448510	Standard
>	Rh	103	13.3	43.3				ug/L	15	Standard
	Mo	98	288.2	19.1	**0.0890**	0.021	23.1	ug/L	11	Standard
	Ag	107	295.0	92.5	**0.0287**	0.042	146.5	ug/L	134	Standard
	Cd	111	58.8	115.8	**0.0253**	0.033	130.1	mg/L	7	Standard
	Cd	114	179.7	150.5	**0.0306**	0.054	175.0	ug/L	16	Standard
>	In	115	602649.0	1.7				ug/L	611117	Standard
	Sn	118	880.0	23.1	**0.0494**	0.037	75.1	ug/L	518	Standard
	Sb	123	564.4	53.6	**0.1091**	0.067	61.4	ug/L	127	Standard
	Ba	135	113.7	104.6	**0.0312**	0.059	190.7	ug/L	45	Standard
	Ce	140	26.7	21.7				ug/L	298	Standard
>	Tb	159	956287.2	0.6				ug/L	977942	Standard
	Ho	165	6.7	173.2				ug/L	12	Standard
	Tl	203	302.3	107.0	**0.0307**	0.038	123.1	ug/L	31	Standard
	Tl	205	353.3	119.4	**0.0454**	0.057	124.8	ug/L	20	Standard
	Pb	206	425.3	47.1	**0.0143**	0.036	251.9	ug/L	286	Standard
	Pb	207	353.0	32.9	**0.0137**	0.023	168.1	ug/L	257	Standard
	Pb	208	1674.0	55.6	**0.0274**	0.048	175.9	ug/L	956	Standard
	U	238	298.0	132.6	**0.0381**	0.054	142.5	ug/L	17	Standard
>	Bi	209	561401.5	0.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:15:18
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	206.7	141.8	**0.2403**	0.424	176.4	mg/L	32	Standard
K	39	28.3	71.3	**-0.0379**	0.076	200.2	mg/L	30	Standard
Ca	43	35.0	24.7	**0.2817**	1.299	461.1	mg/L	48	Standard
Fe	54	147.7	17.2	**0.0057**	0.030	525.6	mg/L	159	Standard
Fe	57	195.0	14.3	**0.0744**	0.099	133.6	mg/L	195	Standard
Sc-1	45	37247.5	4.3				mg/L	39543	Standard
Cl	35	35032.5	2.2				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1893.5	11.2				ug/L	1647	Standard
P	31	18479.6	2.3				ug/L	15231	Standard
S	34	2273.5	3.9				ug/L	1845	Standard
Sr	88	81.7	3.5				ug/L	70	Standard
C	12	186.7	15.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	9.5	186.3				mg/L	12	Standard
Ho-1	165	6.7	173.2				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	2043.5	19.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		97.768	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:15:18
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.614
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	100.776
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:15:18
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115501
Sample Date/Time: Tuesday, June 28, 2016 17:16:14
Number of Replicates: 3
Autosampler Position: 252
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92590.4	1.6				ug/L	81760	Standard
	Be	9	26.7	47.2	**0.0141**	0.012	82.4	ug/L	17	Standard
	Al	27	22123057.0	1.6	**126.7281**	1.652	1.3	ug/L	1410	Standard
	Sc	45	39028.5	1.1				ug/L	39543	Standard
	Ti	47	160.7	8.5	**0.5119**	0.051	9.9	ug/L	39	Standard
	V	51	4388.8	2.7	**0.4554**	0.016	3.5	ug/L	1415	Standard
	Cr	52	10159.7	1.1	**0.4737**	0.053	11.2	ug/L	6946	Standard
	Cr	53	1711.8	8.9	**1.7064**	0.233	13.7	ug/L	328	Standard
	Mn	55	4816073.5	0.9	**779.0277**	27.654	3.5	ug/L	823	Standard
	Co	59	2714.6	0.9	**0.3738**	0.009	2.5	ug/L	252	Standard
	Ni	60	3772.1	1.1	**1.9951**	0.039	2.0	ug/L	192	Standard
	Cu	65	754.7	3.1	**0.3255**	0.008	2.4	ug/L	141	Standard
	Zn	66	4017.9	2.4	**3.0152**	0.057	1.9	ug/L	491	Standard
>	**Ge**	72	491414.0	2.7				ug/L	515087	Standard
	As	75	1051.2	2.3	**1.0210**	0.045	4.4	ug/L	20	Standard
	Se	82	143.6	4.5	**1.1879**	0.030	2.5	ug/L	15	Standard
	Se-1	77	105.7	15.2	**0.7547**	0.249	32.9	ug/L	48	Standard
>	**Ga**	71	188.3	17.7				mg/L	90	Standard
	Rb	85	950.0	0.9				ug/L	23	Standard
	Y	89	424439.8	0.9				ug/L	448510	Standard
>	**Rh**	103	96.7	7.9				ug/L	15	Standard
	Mo	98	5849.5	0.5	**2.2126**	0.029	1.3	ug/L	11	Standard
	Ag	107	89.7	21.1	**-0.0024**	0.003	128.8	ug/L	134	Standard
	Cd	111	11.8	58.9	**0.0028**	0.003	121.4	mg/L	7	Standard
	Cd	114	46.1	44.3	**0.0042**	0.004	96.2	ug/L	16	Standard
>	**In**	115	592706.9	1.7				ug/L	611117	Standard
	Sn	118	891.7	9.5	**0.0535**	0.013	23.6	ug/L	518	Standard
	Sb	123	236.5	14.9	**0.0388**	0.007	18.8	ug/L	127	Standard
	Ba	135	171509.6	0.5	**85.6510**	1.837	2.1	ug/L	45	Standard
	Ce	140	703.3	3.9				ug/L	298	Standard
>	**Tb**	159	954495.6	1.0				ug/L	977942	Standard
	Ho	165	26.7	60.3				ug/L	12	Standard
	Tl	203	128.7	6.5	**0.0108**	0.001	9.4	ug/L	31	Standard
	Tl	205	116.7	2.5	**0.0140**	0.000	2.9	ug/L	20	Standard
	Pb	206	518.0	5.9	**0.0341**	0.006	17.0	ug/L	286	Standard
	Pb	207	434.7	5.8	**0.0328**	0.005	16.1	ug/L	257	Standard
	Pb	208	1843.7	3.4	**0.0392**	0.003	8.9	ug/L	956	Standard
	U	238	11694.2	0.7	**1.6602**	0.012	0.7	ug/L	17	Standard
>	**Bi**	209	546693.4	0.1				ug/L	557080	Standard

Sample ID: L1606115501
Report Date/Time: Tuesday, June 28, 2016 17:18:31
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.9680**	18.988	173.1	mg/L	0	Standard
Mg	24	30276.9	1.0	**40.7793**	0.477	1.2	mg/L	32	Standard
K	39	240.0	12.5	**0.6798**	0.099	14.5	mg/L	30	Standard
Ca	43	255.0	15.3	**33.9356**	6.415	18.9	mg/L	48	Standard
Fe	54	184.3	20.5	**0.0318**	0.034	107.2	mg/L	159	Standard
Fe	57	341.7	3.0	**0.5441**	0.035	6.5	mg/L	195	Standard
> **Sc-1**	45	39028.5	1.1				mg/L	39543	Standard
Cl	35	37440.9	1.2				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	32190.8	1.7				ug/L	1647	Standard
P	31	22541.8	0.7				ug/L	15231	Standard
S	34	2386.9	3.2				ug/L	1845	Standard
Sr	88	303.3	20.6				ug/L	70	Standard
C	12	480.0	6.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	27.6	75.9				mg/L	12	Standard
Ho-1	165	26.7	60.3				mg/L	12	Standard
Er	166	50.0	40.0				mg/L	20	Standard
I	127	1264643.9	6.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.247	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.404	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115501
Report Date/Time: Tuesday, June 28, 2016 17:18:31
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	96.988
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	98.135
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115502
Sample Date/Time: Tuesday, June 28, 2016 17:19:26
Number of Replicates: 3
Autosampler Position: 253
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94157.1	4.7				ug/L	81760	Standard
	Be	9	20.0	50.0	**0.0078**	0.009	115.3	ug/L	17	Standard
	Al	27	28931934.4	4.0	**163.0048**	1.392	0.9	ug/L	1410	Standard
	Sc	45	40320.3	3.5				ug/L	39543	Standard
	Ti	47	168.3	17.1	**0.5232**	0.117	22.4	ug/L	39	Standard
	V	51	1551.4	10.5	**0.0205**	0.027	130.5	ug/L	1415	Standard
	Cr	52	10328.9	2.0	**0.4515**	0.024	5.3	ug/L	6946	Standard
	Cr	53	2665.2	7.9	**2.7435**	0.199	7.2	ug/L	328	Standard
	Mn	55	1534029.1	1.0	**240.4438**	3.129	1.3	ug/L	823	Standard
	Co	59	1102.0	2.8	**0.1281**	0.004	3.4	ug/L	252	Standard
	Ni	60	1786.8	2.8	**0.8597**	0.029	3.3	ug/L	192	Standard
	Cu	65	955.7	1.4	**0.4206**	0.012	3.0	ug/L	141	Standard
	Zn	66	3611.4	2.8	**2.5229**	0.115	4.5	ug/L	491	Standard
>	**Ge**	72	506659.2	1.8				ug/L	515087	Standard
	As	75	1883.3	1.6	**1.7433**	0.024	1.4	ug/L	20	Standard
	Se	82	155.4	0.8	**1.2536**	0.030	2.4	ug/L	15	Standard
	Se-1	77	168.7	6.6	**1.5590**	0.181	11.6	ug/L	48	Standard
>	**Ga**	71	140.0	15.6				mg/L	90	Standard
	Rb	85	1435.1	8.2				ug/L	23	Standard
	Y	89	449200.8	1.5				ug/L	448510	Standard
>	**Rh**	103	93.3	16.4				ug/L	15	Standard
	Mo	98	2699.1	0.6	**0.9667**	0.027	2.8	ug/L	11	Standard
	Ag	107	94.7	12.7	**-0.0023**	0.002	96.2	ug/L	134	Standard
	Cd	111	15.7	16.0	**0.0044**	0.001	31.4	mg/L	7	Standard
	Cd	114	28.9	55.3	**0.0006**	0.003	513.8	ug/L	16	Standard
>	**In**	115	619397.8	3.1				ug/L	611117	Standard
	Sn	118	1035.0	8.9	**0.0707**	0.012	17.4	ug/L	518	Standard
	Sb	123	177.5	18.8	**0.0244**	0.008	32.9	ug/L	127	Standard
	Ba	135	342505.9	1.5	**163.7121**	2.675	1.6	ug/L	45	Standard
	Ce	140	470.0	1.8				ug/L	298	Standard
>	**Tb**	159	978535.3	2.3				ug/L	977942	Standard
	Ho	165	33.3	17.3				ug/L	12	Standard
	Tl	203	95.7	2.4	**0.0067**	0.000	2.7	ug/L	31	Standard
	Tl	205	98.3	7.8	**0.0112**	0.001	8.1	ug/L	20	Standard
	Pb	206	616.7	3.3	**0.0513**	0.003	5.5	ug/L	286	Standard
	Pb	207	520.0	3.8	**0.0494**	0.003	5.9	ug/L	257	Standard
	Pb	208	2224.4	1.1	**0.0583**	0.001	1.4	ug/L	956	Standard
	U	238	39.7	29.5	**0.0026**	0.002	61.0	ug/L	17	Standard
>	**Bi**	209	553220.4	1.1				ug/L	557080	Standard

Sample ID: L1606115502
Report Date/Time: Tuesday, June 28, 2016 17:21:43
Page 1

Approved: June 29, 2016

Na	23	5.0	0.0	**31.5109**	1.087	3.4	mg/L	0	Standard
Mg	24	25140.9	3.4	**32.7721**	0.865	2.6	mg/L	32	Standard
K	39	73.3	23.9	**0.1023**	0.063	61.8	mg/L	30	Standard
Ca	43	243.3	6.6	**30.9203**	2.535	8.2	mg/L	48	Standard
Fe	54	2629.1	2.7	**2.1972**	0.059	2.7	mg/L	159	Standard
Fe	57	1076.7	7.2	**2.9489**	0.380	12.9	mg/L	195	Standard
> **Sc-1**	45	40320.3	3.5				mg/L	39543	Standard
Cl	35	39998.8	2.6				ug/L	27642	Standard
Kr	83	2.7	43.3				ug/L	2	Standard
Br	81	32989.2	3.3				ug/L	1647	Standard
P	31	24089.1	1.8				ug/L	15231	Standard
S	34	2583.6	5.0				ug/L	1845	Standard
Sr	88	433.3	4.7				ug/L	70	Standard
C	12	350.0	5.7				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	25.7	20.9				mg/L	12	Standard
Ho-1	165	33.3	17.3				mg/L	12	Standard
Er	166	20.0	50.0				mg/L	20	Standard
I	127	377170.1	0.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		115.163	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.364	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115502
Report Date/Time: Tuesday, June 28, 2016 17:21:43
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	101.355
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	99.307
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606115502
Report Date/Time: Tuesday, June 28, 2016 17:21:43
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115503
Sample Date/Time: Tuesday, June 28, 2016 17:22:37
Number of Replicates: 3
Autosampler Position: 254
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93460.7	0.5				ug/L	81760	Standard
	Be	9	35.0	14.3	**0.0211**	0.004	21.2	ug/L	17	Standard
	Al	27	14187081.3	1.7	**80.5012**	1.275	1.6	ug/L	1410	Standard
	Sc	45	39556.6	4.7				ug/L	39543	Standard
	Ti	47	229.0	18.7	**0.7435**	0.168	22.5	ug/L	39	Standard
	V	51	2172.0	8.1	**0.1016**	0.022	21.3	ug/L	1415	Standard
	Cr	52	10167.4	1.5	**0.3846**	0.034	8.8	ug/L	6946	Standard
	Cr	53	1090.0	15.2	**0.8840**	0.185	20.9	ug/L	328	Standard
	Mn	55	1432250.3	1.3	**217.9180**	5.450	2.5	ug/L	823	Standard
	Co	59	795.7	6.2	**0.0804**	0.008	9.7	ug/L	252	Standard
	Ni	60	1574.1	4.4	**0.7198**	0.033	4.6	ug/L	192	Standard
	Cu	65	1999.1	3.4	**0.9475**	0.036	3.8	ug/L	141	Standard
	Zn	66	5797.8	1.6	**4.3927**	0.080	1.8	ug/L	491	Standard
>	**Ge**	72	521945.3	1.2				ug/L	515087	Standard
	As	75	168.7	6.2	**0.1883**	0.009	4.6	ug/L	20	Standard
	Se	82	81.0	7.2	**0.5749**	0.055	9.6	ug/L	15	Standard
	Se-1	77	72.3	14.9	**0.2276**	0.131	57.6	ug/L	48	Standard
>	**Ga**	71	155.0	23.3				mg/L	90	Standard
	Rb	85	10031.7	2.6				ug/L	23	Standard
	Y	89	435099.7	2.5				ug/L	448510	Standard
>	**Rh**	103	70.0	28.6				ug/L	15	Standard
	Mo	98	145.2	5.2	**0.0354**	0.003	8.7	ug/L	11	Standard
	Ag	107	85.3	8.5	**-0.0033**	0.001	30.7	ug/L	134	Standard
	Cd	111	76.7	2.3	**0.0335**	0.001	2.5	mg/L	7	Standard
	Cd	114	203.1	19.3	**0.0347**	0.008	22.4	ug/L	16	Standard
>	**In**	115	602973.5	0.6				ug/L	611117	Standard
	Sn	118	941.7	4.9	**0.0595**	0.009	14.6	ug/L	518	Standard
	Sb	123	323.8	6.4	**0.0567**	0.004	7.5	ug/L	127	Standard
	Ba	135	119338.2	0.9	**58.5639**	0.899	1.5	ug/L	45	Standard
	Ce	140	855.0	10.9				ug/L	298	Standard
>	**Tb**	159	952983.8	1.5				ug/L	977942	Standard
	Ho	165	36.7	41.7				ug/L	12	Standard
	Tl	203	88.7	6.9	**0.0057**	0.001	14.0	ug/L	31	Standard
	Tl	205	55.0	27.3	**0.0052**	0.002	38.2	ug/L	20	Standard
	Pb	206	861.0	5.9	**0.0956**	0.010	10.1	ug/L	286	Standard
	Pb	207	708.7	5.0	**0.0873**	0.008	9.7	ug/L	257	Standard
	Pb	208	2964.1	2.3	**0.0967**	0.005	5.3	ug/L	956	Standard
	U	238	696.7	2.0	**0.0942**	0.003	2.9	ug/L	17	Standard
>	**Bi**	209	557355.1	1.0				ug/L	557080	Standard

Sample ID: L1606115503
Report Date/Time: Tuesday, June 28, 2016 17:24:54
Page 1

Approved: June 29, 2016

Na	23	6.7	43.3	**42.2849**	16.098	38.1	mg/L	0	Standard
Mg	24	19544.3	4.8	**25.9961**	1.909	7.3	mg/L	32	Standard
K	39	113.3	15.5	**0.2405**	0.047	19.5	mg/L	30	Standard
Ca	43	286.7	25.9	**38.5627**	12.897	33.4	mg/L	48	Standard
Fe	54	997.4	5.8	**0.7664**	0.064	8.4	mg/L	159	Standard
Fe	57	530.0	13.9	**1.1674**	0.275	23.5	mg/L	195	Standard
> **Sc-1**	45	39556.6	4.7				mg/L	39543	Standard
Cl	35	38409.3	3.3				ug/L	27642	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	17249.8	6.1				ug/L	1647	Standard
P	31	22414.9	3.0				ug/L	15231	Standard
S	34	2625.2	3.4				ug/L	1845	Standard
Sr	88	228.3	6.7				ug/L	70	Standard
C	12	260.0	15.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	21.4	67.1				mg/L	12	Standard
Ho-1	165	36.7	41.7				mg/L	12	Standard
Er	166	40.0	75.0				mg/L	20	Standard
I	127	663837.8	8.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		114.311	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.332	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115503
Report Date/Time: Tuesday, June 28, 2016 17:24:54
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.668
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.049
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606115503
Report Date/Time: Tuesday, June 28, 2016 17:24:54
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 17:25:50
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	91554.4	4.0				ug/L	81760	Standard
	Be	9	59302.5	0.5	**53.3655**	1.843	3.5	ug/L	17	Standard
	Al	27	8323459.6	1.6	**48.2773**	2.634	5.5	ug/L	1410	Standard
	Sc	45	40460.6	1.5				ug/L	39543	Standard
	Ti	47	25609.3	2.1	**100.1533**	1.793	1.8	ug/L	39	Standard
	V	51	364470.3	1.4	**50.8781**	0.853	1.7	ug/L	1415	Standard
	Cr	52	366105.1	1.7	**50.1771**	0.675	1.3	ug/L	6946	Standard
	Cr	53	45260.8	0.4	**49.8893**	0.509	1.0	ug/L	328	Standard
	Mn	55	332731.4	2.3	**49.8507**	1.205	2.4	ug/L	823	Standard
	Co	59	382266.4	2.9	**52.9683**	1.437	2.7	ug/L	252	Standard
	Ni	60	97387.3	1.2	**50.2694**	0.545	1.1	ug/L	192	Standard
	Cu	65	98831.3	1.3	**50.5786**	0.557	1.1	ug/L	141	Standard
	Zn	66	57834.9	1.6	**50.5762**	0.877	1.7	ug/L	491	Standard
>	Ge	72	528768.9	1.2				ug/L	515087	Standard
	As	75	58724.5	1.2	**50.9203**	0.224	0.4	ug/L	20	Standard
	Se	82	6151.0	1.0	**51.9261**	0.245	0.5	ug/L	15	Standard
	Se-1	77	4089.2	2.2	**52.1522**	0.563	1.1	ug/L	48	Standard
>	Ga	71	85.0	21.2				mg/L	90	Standard
	Rb	85	851.7	6.5				ug/L	23	Standard
	Y	89	456324.8	0.4				ug/L	448510	Standard
>	Rh	103	36.7	39.4				ug/L	15	Standard
	Mo	98	284202.4	1.9	**102.0947**	3.648	3.6	ug/L	11	Standard
	Ag	107	351396.4	1.1	**50.8197**	1.435	2.8	ug/L	134	Standard
	Cd	111	110732.8	1.5	**50.3025**	1.470	2.9	mg/L	7	Standard
	Cd	114	271688.9	2.3	**50.7793**	1.915	3.8	ug/L	16	Standard
>	In	115	629514.4	1.7				ug/L	611117	Standard
	Sn	118	310739.2	0.7	**50.9245**	0.924	1.8	ug/L	518	Standard
	Sb	123	246166.6	1.6	**50.6553**	1.659	3.3	ug/L	127	Standard
	Ba	135	104825.0	0.9	**49.2805**	1.265	2.6	ug/L	45	Standard
	Ce	140	70.0	55.8				ug/L	298	Standard
>	Tb	159	987475.2	1.7				ug/L	977942	Standard
	Ho	165	5.0	100.0				ug/L	12	Standard
	Tl	203	444397.7	1.4	**50.9180**	1.232	2.4	ug/L	31	Standard
	Tl	205	396866.3	1.6	**52.3722**	1.695	3.2	ug/L	20	Standard
	Pb	206	284750.7	1.1	**50.8069**	1.281	2.5	ug/L	286	Standard
	Pb	207	259280.5	1.3	**51.2940**	1.414	2.8	ug/L	257	Standard
	Pb	208	1040337.0	1.7	**53.1998**	1.656	3.1	ug/L	956	Standard
	U	238	390662.1	1.3	**52.6405**	1.349	2.6	ug/L	17	Standard
>	Bi	209	577126.2	1.6				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:28:07
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4445.7	6.0	**5.7289**	0.397	6.9	mg/L	32	Standard
K	39	1320.1	5.5	**4.2194**	0.276	6.5	mg/L	30	Standard
Ca	43	81.7	24.7	**6.7579**	2.904	43.0	mg/L	48	Standard
Fe	54	5927.1	5.8	**5.1036**	0.243	4.8	mg/L	159	Standard
Fe	57	1661.8	3.7	**4.8569**	0.121	2.5	mg/L	195	Standard
Sc-1	45	40460.6	1.5				mg/L	39543	Standard
Cl	35	38764.2	1.7				ug/L	27642	Standard
Kr	83	2.3	107.9				ug/L	2	Standard
Br	81	2123.5	9.4				ug/L	1647	Standard
P	31	21915.8	3.2				ug/L	15231	Standard
S	34	2866.9	6.9				ug/L	1845	Standard
Sr	88	56.7	10.2				ug/L	70	Standard
C	12	170.0	44.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	173.2				mg/L	0	Standard
Dy	164	9.7	103.3				mg/L	12	Standard
Ho-1	165	5.0	100.0				mg/L	12	Standard
Er	166	6.7	86.6				mg/L	20	Standard
I	127	21360.9	50.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	106.731		
Al	27	96.555		
Sc	45			
Ti	47	100.153		
V	51	101.756		
Cr	52	100.354		
Cr	53			
Mn	55	99.701		
Co	59	105.937		
Ni	60	100.539		
Cu	65	101.157		
Zn	66	101.152		
Ge	72		102.656	
As	75	101.841		
Se	82	103.852		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:28:07
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	102.095		
Ag	107	101.639		
Cd	111	100.605		
Cd	114			
> In	115			103.011
Sn	118	101.849		
Sb	123	101.311		
Ba	135	98.561		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.836		
Tl	205			
Pb	206			
Pb	207			
Pb	208	106.400		
U	238	105.281		
> Bi	209			103.598
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:28:07
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 17:29:02
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89289.0	1.1				ug/L	81760	Standard
	Be	9	35.0	51.5	**0.0225**	0.017	74.1	ug/L	17	Standard
	Al	27	14272.5	142.7	**0.0706**	0.122	172.0	ug/L	1410	Standard
	Sc	45	39427.9	1.8				ug/L	39543	Standard
	Ti	47	28.0	28.3	**-0.0534**	0.032	59.3	ug/L	39	Standard
	V	51	1205.0	11.4	**-0.0346**	0.020	56.4	ug/L	1415	Standard
	Cr	52	6173.6	1.1	**-0.1773**	0.009	5.0	ug/L	6946	Standard
	Cr	53	500.0	6.0	**0.2236**	0.034	15.3	ug/L	328	Standard
	Mn	55	1435.1	83.0	**0.0711**	0.182	256.2	ug/L	823	Standard
	Co	59	252.7	26.0	**0.0043**	0.009	214.7	ug/L	252	Standard
	Ni	60	265.7	8.4	**0.0350**	0.012	33.5	ug/L	192	Standard
	Cu	65	144.0	13.3	**-0.0157**	0.010	63.6	ug/L	141	Standard
	Zn	66	428.7	4.3	**-0.4396**	0.017	3.8	ug/L	491	Standard
>	**Ge**	72	519199.6	0.1				ug/L	515087	Standard
	As	75	-13.2	109.3	**0.0287**	0.013	44.2	ug/L	20	Standard
	Se	82	24.7	35.1	**0.0933**	0.075	79.8	ug/L	15	Standard
	Se-1	77	53.0	16.4	**-0.0209**	0.115	551.4	ug/L	48	Standard
>	**Ga**	71	56.7	41.7				mg/L	90	Standard
	Rb	85	13.3	78.1				ug/L	23	Standard
	Y	89	435848.4	0.6				ug/L	448510	Standard
>	**Rh**	103	16.7	45.8				ug/L	15	Standard
	Mo	98	252.5	21.9	**0.0755**	0.022	29.1	ug/L	11	Standard
	Ag	107	123.0	3.3	**0.0023**	0.000	14.7	ug/L	134	Standard
	Cd	111	11.8	114.0	**0.0027**	0.006	236.7	mg/L	7	Standard
	Cd	114	47.6	83.9	**0.0044**	0.008	178.8	ug/L	16	Standard
>	**In**	115	604937.7	1.9				ug/L	611117	Standard
	Sn	118	835.0	26.8	**0.0403**	0.035	87.6	ug/L	518	Standard
	Sb	123	438.3	28.1	**0.0807**	0.025	30.6	ug/L	127	Standard
	Ba	135	40.7	30.0	**-0.0052**	0.006	117.6	ug/L	45	Standard
	Ce	140	35.0	37.8				ug/L	298	Standard
>	**Tb**	159	968167.7	1.8				ug/L	977942	Standard
	Ho	165	8.3	124.9				ug/L	12	Standard
	Tl	203	65.7	40.9	**0.0028**	0.003	111.1	ug/L	31	Standard
	Tl	205	198.3	119.2	**0.0243**	0.032	131.1	ug/L	20	Standard
	Pb	206	312.0	4.9	**-0.0071**	0.003	48.0	ug/L	286	Standard
	Pb	207	265.7	2.9	**-0.0047**	0.002	47.8	ug/L	257	Standard
	Pb	208	1274.7	38.3	**0.0057**	0.026	450.9	ug/L	956	Standard
	U	238	172.3	130.6	**0.0207**	0.031	149.9	ug/L	17	Standard
>	**Bi**	209	569803.2	1.6				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:31:19
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**21.6890**	18.779	86.6	mg/L	0	Standard
Mg	24	61.7	47.5	**0.0256**	0.040	158.0	mg/L	32	Standard
K	39	26.7	10.8	**-0.0510**	0.010	20.5	mg/L	30	Standard
Ca	43	21.7	48.0	**-2.0553**	1.552	75.5	mg/L	48	Standard
Fe	54	129.4	21.4	**-0.0198**	0.024	119.8	mg/L	159	Standard
Fe	57	196.7	34.5	**0.0395**	0.221	559.0	mg/L	195	Standard
Sc-1	45	39427.9	1.8				mg/L	39543	Standard
Cl	35	37196.3	1.0				ug/L	27642	Standard
Kr	83	2.7	21.7				ug/L	2	Standard
Br	81	1853.4	7.9				ug/L	1647	Standard
P	31	20680.8	2.7				ug/L	15231	Standard
S	34	2700.2	3.5				ug/L	1845	Standard
Sr	88	68.3	23.5				ug/L	70	Standard
C	12	133.3	38.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.5	107.6				mg/L	12	Standard
Ho-1	165	8.3	124.9				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	7115.0	15.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		100.798	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:31:19
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	98.989	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	102.284	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:31:19
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606114201
Sample Date/Time: Tuesday, June 28, 2016 18:46:32
Number of Replicates: 3
Autosampler Position: 255
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78164.5	2.4				ug/L	81760	Standard
	Be	9	18.3	41.7	**0.0097**	0.008	86.8	ug/L	17	Standard
	Al	27	121659.9	1.7	**0.8116**	0.028	3.5	ug/L	1410	Standard
	Sc	45	40555.9	2.8				ug/L	39543	Standard
	Ti	47	55.7	15.0	**0.0583**	0.036	61.2	ug/L	39	Standard
	V	51	1460.1	6.7	**0.0028**	0.012	438.4	ug/L	1415	Standard
	Cr	52	7975.8	0.8	**0.0855**	0.018	21.3	ug/L	6946	Standard
	Cr	53	690.0	12.1	**0.4421**	0.091	20.5	ug/L	328	Standard
	Mn	55	73452.6	0.4	**11.1499**	0.088	0.8	ug/L	823	Standard
	Co	59	2365.9	3.2	**0.3044**	0.010	3.2	ug/L	252	Standard
	Ni	60	956.0	2.8	**0.4012**	0.010	2.4	ug/L	192	Standard
	Cu	65	203.7	5.3	**0.0160**	0.005	33.1	ug/L	141	Standard
	Zn	66	1994.1	0.1	**0.9866**	0.018	1.8	ug/L	491	Standard
>	**Ge**	72	516559.4	1.0				ug/L	515087	Standard
	As	75	-28.9	32.9	**0.0147**	0.008	56.4	ug/L	20	Standard
	Se	82	19.3	15.0	**0.0479**	0.026	53.5	ug/L	15	Standard
	Se-1	77	59.0	13.2	**0.0614**	0.097	157.2	ug/L	48	Standard
>	**Ga**	71	36.7	51.6				mg/L	90	Standard
	Rb	85	463.3	2.5				ug/L	23	Standard
	Y	89	448462.9	2.6				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	65.9	7.4	**0.0053**	0.002	42.6	ug/L	11	Standard
	Ag	107	82.7	4.2	**-0.0039**	0.001	16.6	ug/L	134	Standard
	Cd	111	8.9	29.8	**0.0012**	0.001	94.7	mg/L	7	Standard
	Cd	114	20.4	73.3	**-0.0010**	0.003	278.6	ug/L	16	Standard
>	**In**	115	611220.6	1.9				ug/L	611117	Standard
	Sn	118	670.0	14.0	**0.0115**	0.017	148.1	ug/L	518	Standard
	Sb	123	214.2	30.6	**0.0326**	0.014	43.4	ug/L	127	Standard
	Ba	135	7677.9	3.5	**3.6947**	0.165	4.5	ug/L	45	Standard
	Ce	140	396.7	10.7				ug/L	298	Standard
>	**Tb**	159	963569.5	0.7				ug/L	977942	Standard
	Ho	165	16.7	45.8				ug/L	12	Standard
	Tl	203	21.7	17.5	**-0.0022**	0.000	19.5	ug/L	31	Standard
	Tl	205	13.3	21.7	**-0.0005**	0.000	79.8	ug/L	20	Standard
	Pb	206	368.7	4.6	**0.0040**	0.003	86.3	ug/L	286	Standard
	Pb	207	283.0	4.1	**-0.0005**	0.002	459.4	ug/L	257	Standard
	Pb	208	1139.7	3.0	**-0.0006**	0.002	246.5	ug/L	956	Standard
	U	238	10.3	39.1	**-0.0015**	0.001	36.3	ug/L	17	Standard
>	**Bi**	209	562456.7	0.5				ug/L	557080	Standard

Sample ID: L1606114201
Report Date/Time: Tuesday, June 28, 2016 18:48:49
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	763.4	13.7	**0.9351**	0.146	15.6	mg/L	32	Standard
K	39	31.7	39.7	**-0.0363**	0.045	123.6	mg/L	30	Standard
Ca	43	18.3	31.5	**-2.6404**	0.796	30.1	mg/L	48	Standard
Fe	54	359.1	17.1	**0.1792**	0.047	26.3	mg/L	159	Standard
Fe	57	273.3	13.7	**0.2778**	0.149	53.6	mg/L	195	Standard
> **Sc-1**	45	40555.9	2.8				mg/L	39543	Standard
Cl	35	37427.5	0.7				ug/L	27642	Standard
Kr	83	0.0					ug/L	2	Standard
Br	81	2536.9	6.5				ug/L	1647	Standard
P	31	19003.6	0.9				ug/L	15231	Standard
S	34	2375.2	4.8				ug/L	1845	Standard
Sr	88	91.7	13.7				ug/L	70	Standard
C	12	86.7	29.0				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	29.2	34.3				mg/L	12	Standard
Ho-1	165	16.7	45.8				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	22815.7	12.2				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		95.602	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		100.286	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606114201
Report Date/Time: Tuesday, June 28, 2016 18:48:49
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.017
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.965
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 18:49:46
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	77476.0	3.1				ug/L	81760	Standard
	Be	9	45015.2	3.8	**47.8878**	3.193	6.7	ug/L	17	Standard
	Al	27	6953697.9	2.2	**47.6359**	2.288	4.8	ug/L	1410	Standard
	Sc	45	38896.5	2.1				ug/L	39543	Standard
	Ti	47	23530.3	2.9	**93.4171**	3.639	3.9	ug/L	39	Standard
	V	51	333406.5	0.5	**47.2267**	0.631	1.3	ug/L	1415	Standard
	Cr	52	337441.2	2.6	**46.8861**	1.719	3.7	ug/L	6946	Standard
	Cr	53	42128.5	3.1	**47.1231**	1.874	4.0	ug/L	328	Standard
	Mn	55	311025.1	1.6	**47.2965**	1.233	2.6	ug/L	823	Standard
	Co	59	333541.6	0.3	**46.9106**	0.523	1.1	ug/L	252	Standard
	Ni	60	90766.5	2.7	**47.5598**	1.737	3.7	ug/L	192	Standard
	Cu	65	91363.1	2.5	**47.4629**	1.646	3.5	ug/L	141	Standard
	Zn	66	53283.5	1.9	**47.2487**	1.367	2.9	ug/L	491	Standard
>	Ge	72	520941.4	1.0				ug/L	515087	Standard
	As	75	53453.8	0.5	**47.0530**	0.571	1.2	ug/L	20	Standard
	Se	82	5409.0	0.9	**46.3406**	0.883	1.9	ug/L	15	Standard
	Se-1	77	3695.1	1.5	**47.7773**	0.360	0.8	ug/L	48	Standard
>	Ga	71	58.3	27.6				mg/L	90	Standard
	Rb	85	731.7	6.2				ug/L	23	Standard
	Y	89	433135.5	0.5				ug/L	448510	Standard
>	Rh	103	21.7	13.3				ug/L	15	Standard
	Mo	98	251705.5	0.6	**94.1208**	1.331	1.4	ug/L	11	Standard
	Ag	107	314845.2	1.8	**47.3955**	0.605	1.3	ug/L	134	Standard
	Cd	111	98007.4	2.2	**46.3425**	0.750	1.6	mg/L	7	Standard
	Cd	114	239351.7	3.2	**46.5659**	1.544	3.3	ug/L	16	Standard
>	In	115	604583.1	1.4				ug/L	611117	Standard
	Sn	118	276455.2	2.3	**47.1543**	0.498	1.1	ug/L	518	Standard
	Sb	123	224560.9	2.5	**48.0923**	0.641	1.3	ug/L	127	Standard
	Ba	135	94923.9	3.0	**46.4488**	1.094	2.4	ug/L	45	Standard
	Ce	140	80.0	22.5				ug/L	298	Standard
>	Tb	159	958457.8	2.2				ug/L	977942	Standard
	Ho	165	13.3	21.7				ug/L	12	Standard
	Tl	203	399564.4	2.5	**47.4112**	1.332	2.8	ug/L	31	Standard
	Tl	205	347775.3	2.0	**47.5216**	1.083	2.3	ug/L	20	Standard
	Pb	206	259967.8	1.3	**48.0301**	0.722	1.5	ug/L	286	Standard
	Pb	207	236539.5	0.9	**48.4530**	0.488	1.0	ug/L	257	Standard
	Pb	208	913309.7	2.1	**48.3569**	1.176	2.4	ug/L	956	Standard
	U	238	337839.3	1.6	**47.1413**	0.910	1.9	ug/L	17	Standard
>	Bi	209	557186.5	0.3				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 18:52:03
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3633.8	7.1	**4.8666**	0.455	9.3	mg/L	32	Standard
K	39	1313.4	12.2	**4.3752**	0.599	13.7	mg/L	30	Standard
Ca	43	51.7	20.1	**2.6371**	1.625	61.6	mg/L	48	Standard
Fe	54	5289.1	7.8	**4.7357**	0.483	10.2	mg/L	159	Standard
Fe	57	1531.7	10.1	**4.6389**	0.613	13.2	mg/L	195	Standard
> **Sc-1**	45	38896.5	2.1				mg/L	39543	Standard
Cl	35	35022.4	1.4				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	1490.1	8.7				ug/L	1647	Standard
P	31	19075.3	3.9				ug/L	15231	Standard
S	34	2581.9	1.6				ug/L	1845	Standard
Sr	88	70.0	18.9				ug/L	70	Standard
C	12	143.3	33.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	6.0	193.9				mg/L	12	Standard
Ho-1	165	13.3	21.7				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	2193.5	28.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	95.776		
Al	27	95.272		
Sc	45			
Ti	47	93.417		
V	51	94.453		
Cr	52	93.772		
Cr	53			
Mn	55	94.593		
Co	59	93.821		
Ni	60	95.120		
Cu	65	94.926		
Zn	66	94.497		
> Ge	72		101.137	
As	75	94.106		
Se	82	92.681		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 18:52:03
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	94.121		
Ag	107	94.791		
Cd	111	92.685		
Cd	114			
> In	115			98.931
Sn	118	94.309		
Sb	123	96.185		
Ba	135	92.898		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	94.822		
Tl	205			
Pb	206			
Pb	207			
Pb	208	96.714		
U	238	94.283		
> Bi	209			100.019
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 18:52:03
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 18:52:57
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75264.9	3.1				ug/L	81760	Standard
	Be	9	28.3	44.4	**0.0211**	0.013	62.2	ug/L	17	Standard
	Al	27	1848.4	7.3	**-0.0014**	0.001	57.2	ug/L	1410	Standard
	Sc	45	39733.8	3.9				ug/L	39543	Standard
	Ti	47	30.7	21.2	**-0.0408**	0.026	64.0	ug/L	39	Standard
	V	51	1157.9	3.5	**-0.0386**	0.007	18.1	ug/L	1415	Standard
	Cr	52	6209.6	2.0	**-0.1578**	0.015	9.7	ug/L	6946	Standard
	Cr	53	373.3	6.7	**0.0873**	0.028	32.1	ug/L	328	Standard
	Mn	55	647.3	1.5	**-0.0478**	0.002	4.7	ug/L	823	Standard
	Co	59	232.0	9.1	**0.0019**	0.003	163.0	ug/L	252	Standard
	Ni	60	166.3	9.0	**-0.0160**	0.007	44.1	ug/L	192	Standard
	Cu	65	124.7	14.0	**-0.0248**	0.008	34.1	ug/L	141	Standard
	Zn	66	405.0	8.1	**-0.4552**	0.029	6.3	ug/L	491	Standard
>	**Ge**	72	510975.0	1.4				ug/L	515087	Standard
	As	75	-43.1	41.4	**0.0017**	0.016	900.9	ug/L	20	Standard
	Se	82	19.6	25.0	**0.0525**	0.045	86.4	ug/L	15	Standard
	Se-1	77	59.7	5.9	**0.0797**	0.050	62.7	ug/L	48	Standard
>	**Ga**	71	36.7	39.4				mg/L	90	Standard
	Rb	85	23.3	32.7				ug/L	23	Standard
	Y	89	435979.5	2.5				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	250.3	25.6	**0.0730**	0.024	32.3	ug/L	11	Standard
	Ag	107	100.0	11.8	**-0.0014**	0.002	125.8	ug/L	134	Standard
	Cd	111	11.9	34.0	**0.0026**	0.002	71.6	mg/L	7	Standard
	Cd	114	45.3	6.7	**0.0038**	0.000	13.0	ug/L	16	Standard
>	**In**	115	614932.8	1.3				ug/L	611117	Standard
	Sn	118	795.0	23.6	**0.0316**	0.031	97.7	ug/L	518	Standard
	Sb	123	431.0	25.7	**0.0779**	0.023	29.6	ug/L	127	Standard
	Ba	135	30.3	24.3	**-0.0105**	0.003	32.1	ug/L	45	Standard
	Ce	140	20.0	66.1				ug/L	298	Standard
>	**Tb**	159	946813.4	1.4				ug/L	977942	Standard
	Ho	165	11.7	89.2				ug/L	12	Standard
	Tl	203	48.3	29.0	**0.0009**	0.002	172.8	ug/L	31	Standard
	Tl	205	35.0	86.9	**0.0025**	0.004	165.9	ug/L	20	Standard
	Pb	206	284.0	5.8	**-0.0111**	0.003	26.2	ug/L	286	Standard
	Pb	207	257.7	11.7	**-0.0053**	0.006	110.9	ug/L	257	Standard
	Pb	208	990.3	3.8	**-0.0080**	0.002	20.6	ug/L	956	Standard
	U	238	43.0	24.2	**0.0030**	0.001	48.0	ug/L	17	Standard
>	**Bi**	209	557905.1	0.6				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 18:55:14
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	31.7	18.2	**-0.0149**	0.009	60.8	mg/L	32	Standard
K	39	20.0	50.0	**-0.0742**	0.034	46.2	mg/L	30	Standard
Ca	43	35.0	62.3	**-0.0114**	3.447	30330.8	mg/L	48	Standard
Fe	54	144.1	21.6	**-0.0065**	0.032	498.8	mg/L	159	Standard
Fe	57	238.3	3.2	**0.1775**	0.054	30.7	mg/L	195	Standard
Sc-1	45	39733.8	3.9				mg/L	39543	Standard
Cl	35	35783.5	1.7				ug/L	27642	Standard
Kr	83	3.3	45.8				ug/L	2	Standard
Br	81	1466.7	6.9				ug/L	1647	Standard
P	31	19612.7	1.4				ug/L	15231	Standard
S	34	2756.9	2.6				ug/L	1845	Standard
Sr	88	66.7	37.0				ug/L	70	Standard
C	12	86.7	6.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	32.7	47.4				mg/L	12	Standard
Ho-1	165	11.7	89.2				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	1593.4	21.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.202	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 18:55:14
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.624
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.148
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 18:55:14
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Tuesday, June 28, 2016 18:56:10
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	71096.5	2.9				ug/L	81760	Standard
	Be	9	180.0	15.5	**0.1982**	0.026	13.3	ug/L	17	Standard
	Al	27	1130.0	11.5	**-0.0060**	0.001	12.9	ug/L	1410	Standard
	Sc	45	37422.8	0.9				ug/L	39543	Standard
	Ti	47	25.0	10.6	**-0.0625**	0.010	16.0	ug/L	39	Standard
	V	51	3508.6	1.4	**0.3092**	0.005	1.6	ug/L	1415	Standard
	Cr	52	11162.1	1.4	**0.5814**	0.031	5.3	ug/L	6946	Standard
	Cr	53	975.0	4.9	**0.7938**	0.066	8.3	ug/L	328	Standard
	Mn	55	3454.1	1.9	**0.3960**	0.015	3.8	ug/L	823	Standard
	Co	59	2721.9	4.1	**0.3646**	0.019	5.1	ug/L	252	Standard
	Ni	60	2839.3	1.0	**1.4356**	0.026	1.8	ug/L	192	Standard
	Cu	65	1518.4	2.5	**0.7258**	0.025	3.5	ug/L	141	Standard
	Zn	66	6972.6	1.8	**5.6720**	0.065	1.1	ug/L	491	Standard
>	**Ge**	72	504116.3	0.9				ug/L	515087	Standard
	As	75	374.9	9.7	**0.3808**	0.031	8.0	ug/L	20	Standard
	Se	82	63.0	8.6	**0.4404**	0.052	11.9	ug/L	15	Standard
	Se-1	77	80.7	10.3	**0.3758**	0.122	32.5	ug/L	48	Standard
>	**Ga**	71	33.3	52.7				mg/L	90	Standard
	Rb	85	16.7	96.4				ug/L	23	Standard
	Y	89	427965.8	1.5				ug/L	448510	Standard
>	**Rh**	103	20.0	25.0				ug/L	15	Standard
	Mo	98	72.3	5.2	**0.0094**	0.001	12.5	ug/L	11	Standard
	Ag	107	2402.2	4.6	**0.3647**	0.015	4.1	ug/L	134	Standard
	Cd	111	440.5	3.4	**0.2165**	0.007	3.2	mg/L	7	Standard
	Cd	114	1158.4	0.8	**0.2324**	0.001	0.2	ug/L	16	Standard
>	**In**	115	574029.4	1.1				ug/L	611117	Standard
	Sn	118	496.7	16.4	**-0.0124**	0.016	125.3	ug/L	518	Standard
	Sb	123	1769.4	0.8	**0.3864**	0.003	0.7	ug/L	127	Standard
	Ba	135	1350.1	0.7	**0.6712**	0.012	1.8	ug/L	45	Standard
	Ce	140	15.0	0.0				ug/L	298	Standard
>	**Tb**	159	914394.0	1.1				ug/L	977942	Standard
	Ho	165	6.7	43.3				ug/L	12	Standard
	Tl	203	645.7	4.5	**0.0752**	0.004	4.8	ug/L	31	Standard
	Tl	205	546.7	2.9	**0.0757**	0.002	3.3	ug/L	20	Standard
	Pb	206	1234.7	1.4	**0.1748**	0.002	1.1	ug/L	286	Standard
	Pb	207	1077.0	1.2	**0.1725**	0.004	2.3	ug/L	257	Standard
	Pb	208	4484.6	3.2	**0.1877**	0.008	4.4	ug/L	956	Standard
	U	238	2602.6	1.7	**0.3761**	0.010	2.6	ug/L	17	Standard
>	**Bi**	209	533901.6	0.9				ug/L	557080	Standard

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 18:58:27
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	26.7	10.8	**-0.0196**	0.004	19.0	mg/L	32	Standard
K	39	28.3	10.2	**-0.0403**	0.010	24.0	mg/L	30	Standard
Ca	43	38.3	7.5	**0.8090**	0.513	63.4	mg/L	48	Standard
Fe	54	99.0	17.7	**-0.0423**	0.018	41.4	mg/L	159	Standard
Fe	57	198.3	16.4	**0.0831**	0.119	142.8	mg/L	195	Standard
Sc-1	45	37422.8	0.9				mg/L	39543	Standard
Cl	35	34311.5	2.9				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1263.4	4.0				ug/L	1647	Standard
P	31	12998.9	1.6				ug/L	15231	Standard
S	34	2533.5	5.7				ug/L	1845	Standard
Sr	88	71.7	29.0				ug/L	70	Standard
C	12	70.0	24.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	5.9	100.8				mg/L	12	Standard
Ho-1	165	6.7	43.3				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	1311.7	25.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	99.121		
Al	27			
Sc	45			
Ti	47			
V	51	77.290		
Cr	52	72.680		
Cr	53			
Mn	55	79.198		
Co	59	91.138		
Ni	60	89.723		
Cu	65	90.724		
Zn	66	90.752		
Ge	72		97.870	
As	75	95.209		
Se	82	110.096		
Se-1	77			
Ga	71			

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 18:58:27
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	91.178		
Cd	111	90.212		
Cd	114			
> In	115			93.931
Sn	118			
Sb	123	96.601		
Ba	135	89.490		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	93.978		
Tl	205			
Pb	206			
Pb	207			
Pb	208	93.825		
U	238	94.023		
> Bi	209			95.839
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 18:58:27
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: PBW A9 WG572242-02
Sample Date/Time: Tuesday, June 28, 2016 19:51:29
Number of Replicates: 3
Autosampler Position: 301
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73456.0	1.1				ug/L	81760	Standard
	Be	9	13.3	21.7	**0.0052**	0.003	59.5	ug/L	17	Standard
	Al	27	9322.9	1.9	**0.0529**	0.001	1.2	ug/L	1410	Standard
	Sc	45	38756.2	2.9				ug/L	39543	Standard
	Ti	47	33.7	15.2	**-0.0290**	0.022	74.1	ug/L	39	Standard
	V	51	1394.3	5.5	**-0.0051**	0.010	198.0	ug/L	1415	Standard
	Cr	52	7883.4	0.8	**0.0807**	0.013	16.5	ug/L	6946	Standard
	Cr	53	506.7	16.0	**0.2387**	0.095	39.9	ug/L	328	Standard
	Mn	55	1210.0	2.0	**0.0390**	0.005	11.7	ug/L	823	Standard
	Co	59	188.0	6.5	**-0.0045**	0.002	42.5	ug/L	252	Standard
	Ni	60	230.3	1.3	**0.0179**	0.002	10.0	ug/L	192	Standard
	Cu	65	157.0	8.9	**-0.0078**	0.008	102.0	ug/L	141	Standard
	Zn	66	1546.4	2.1	**0.5897**	0.023	3.8	ug/L	491	Standard
>	**Ge**	72	512703.5	0.7				ug/L	515087	Standard
	As	75	-32.6	48.2	**0.0111**	0.014	127.3	ug/L	20	Standard
	Se	82	19.2	20.1	**0.0486**	0.034	70.8	ug/L	15	Standard
	Se-1	77	50.0	10.6	**-0.0518**	0.075	145.3	ug/L	48	Standard
>	**Ga**	71	33.3	34.6				mg/L	90	Standard
	Rb	85	33.3	31.2				ug/L	23	Standard
	Y	89	438886.4	1.2				ug/L	448510	Standard
>	**Rh**	103	13.3	43.3				ug/L	15	Standard
	Mo	98	27.2	14.3	**-0.0088**	0.001	17.0	ug/L	11	Standard
	Ag	107	75.7	8.8	**-0.0047**	0.001	18.4	ug/L	134	Standard
	Cd	111	4.3	48.7	**-0.0009**	0.001	113.9	mg/L	7	Standard
	Cd	114	21.5	41.2	**-0.0007**	0.002	246.6	ug/L	16	Standard
>	**In**	115	598021.6	1.3				ug/L	611117	Standard
	Sn	118	688.3	6.3	**0.0170**	0.007	40.8	ug/L	518	Standard
	Sb	123	162.8	37.0	**0.0223**	0.013	56.3	ug/L	127	Standard
	Ba	135	50.0	12.0	**-0.0003**	0.003	1000.2	ug/L	45	Standard
	Ce	140	36.7	78.7				ug/L	298	Standard
>	**Tb**	159	940667.7	2.2				ug/L	977942	Standard
	Ho	165	18.3	56.8				ug/L	12	Standard
	Tl	203	9.0	22.2	**-0.0037**	0.000	6.2	ug/L	31	Standard
	Tl	205	5.0	0.0	**-0.0016**	0.000	0.7	ug/L	20	Standard
	Pb	206	302.0	8.3	**-0.0074**	0.006	75.2	ug/L	286	Standard
	Pb	207	237.7	5.4	**-0.0090**	0.003	32.3	ug/L	257	Standard
	Pb	208	966.3	1.0	**-0.0090**	0.001	6.5	ug/L	956	Standard
	U	238	4.3	48.0	**-0.0024**	0.000	12.3	ug/L	17	Standard
>	**Bi**	209	554162.6	1.6				ug/L	557080	Standard

Sample ID: PBW A9 WG572242-02
Report Date/Time: Tuesday, June 28, 2016 19:53:46
Page 1

Approved: June 29, 2016

	Na	23	0.0		**0.0050**			mg/L	0	Standard
	Mg	24	41.7	18.3	**-0.0005**	0.010	2017.0	mg/L	32	Standard
	K	39	30.0	33.3	**-0.0385**	0.033	84.7	mg/L	30	Standard
	Ca	43	16.7	91.7	**-2.7218**	2.434	89.4	mg/L	48	Standard
	Fe	54	121.1	16.6	**-0.0249**	0.021	84.8	mg/L	159	Standard
	Fe	57	230.0	15.2	**0.1658**	0.100	60.0	mg/L	195	Standard
>	**Sc-1**	45	38756.2	2.9				mg/L	39543	Standard
	Cl	35	36900.9	1.9				ug/L	27642	Standard
	Kr	83	1.3	43.3				ug/L	2	Standard
	Br	81	1656.8	8.0				ug/L	1647	Standard
	P	31	17782.1	2.2				ug/L	15231	Standard
	S	34	2408.5	10.1				ug/L	1845	Standard
	Sr	88	75.0	29.1				ug/L	70	Standard
	C	12	50.0	52.9				mg/L	113	Standard
	N	14	0.0					mg/L	0	Standard
	Hg	202	3.3	173.2				mg/L	0	Standard
	Dy	164	9.5	105.0				mg/L	12	Standard
	Ho-1	165	18.3	56.8				mg/L	12	Standard
	Er	166	10.0					mg/L	20	Standard
	I	127	1248.4	23.8				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		89.843	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		99.537	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBW A9 WG572242-02
Report Date/Time: Tuesday, June 28, 2016 19:53:46
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.857
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.476
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW A9 WG572242-02
Report Date/Time: Tuesday, June 28, 2016 19:53:46
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: LCSW A9 WG572242-03
Sample Date/Time: Tuesday, June 28, 2016 19:54:40
Number of Replicates: 3
Autosampler Position: 302
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74125.9	0.7				ug/L	81760	Standard
	Be	9	45613.6	2.5	**50.6567**	1.536	3.0	ug/L	17	Standard
	Al	27	52265.0	2.7	**0.3596**	0.008	2.3	ug/L	1410	Standard
	Sc	45	39738.7	2.1				ug/L	39543	Standard
	Ti	47	34.0	10.6	**-0.0277**	0.015	52.7	ug/L	39	Standard
	V	51	340531.5	1.6	**49.0319**	0.683	1.4	ug/L	1415	Standard
	Cr	52	354579.7	0.6	**50.1386**	0.494	1.0	ug/L	6946	Standard
	Cr	53	44436.8	3.3	**50.5389**	1.845	3.7	ug/L	328	Standard
	Mn	55	331885.6	0.6	**51.3034**	0.483	0.9	ug/L	823	Standard
	Co	59	341252.7	1.0	**48.7823**	0.497	1.0	ug/L	252	Standard
	Ni	60	94292.8	0.4	**50.2138**	0.089	0.2	ug/L	192	Standard
	Cu	65	96497.7	0.5	**50.9494**	0.130	0.3	ug/L	141	Standard
	Zn	66	57251.0	0.4	**51.6689**	0.417	0.8	ug/L	491	Standard
>	**Ge**	72	512516.5	0.4				ug/L	515087	Standard
	As	75	55356.1	0.6	**49.5225**	0.094	0.2	ug/L	20	Standard
	Se	82	5489.9	0.8	**47.8062**	0.574	1.2	ug/L	15	Standard
	Se-1	77	3699.5	2.3	**48.6318**	0.978	2.0	ug/L	48	Standard
>	**Ga**	71	51.7	29.6				mg/L	90	Standard
	Rb	85	58.3	27.6				ug/L	23	Standard
	Y	89	427399.3	0.3				ug/L	448510	Standard
>	**Rh**	103	20.0	66.1				ug/L	15	Standard
	Mo	98	50.9	5.4	**0.0000**	0.001	4406.4	ug/L	11	Standard
	Ag	107	325451.5	0.2	**49.1242**	0.412	0.8	ug/L	134	Standard
	Cd	111	100950.5	1.5	**47.8617**	0.611	1.3	mg/L	7	Standard
	Cd	114	243935.8	0.5	**47.5806**	0.389	0.8	ug/L	16	Standard
>	**In**	115	602979.8	0.8				ug/L	611117	Standard
	Sn	118	653.3	9.2	**0.0100**	0.010	97.1	ug/L	518	Standard
	Sb	123	224336.4	0.4	**48.1778**	0.461	1.0	ug/L	127	Standard
	Ba	135	98286.4	0.2	**48.2270**	0.385	0.8	ug/L	45	Standard
	Ce	140	205.0	19.4				ug/L	298	Standard
>	**Tb**	159	962904.7	0.9				ug/L	977942	Standard
	Ho	165	11.7	49.5				ug/L	12	Standard
	Tl	203	409811.5	0.9	**48.6592**	0.503	1.0	ug/L	31	Standard
	Tl	205	358097.5	0.9	**48.9662**	0.679	1.4	ug/L	20	Standard
	Pb	206	272798.6	0.1	**50.4414**	0.712	1.4	ug/L	286	Standard
	Pb	207	236678.2	0.4	**48.5210**	0.945	1.9	ug/L	257	Standard
	Pb	208	941837.1	1.1	**49.9011**	0.393	0.8	ug/L	956	Standard
	U	238	343975.4	0.3	**48.0328**	0.703	1.5	ug/L	17	Standard
>	**Bi**	209	556837.1	1.5				ug/L	557080	Standard

Sample ID: LCSW A9 WG572242-03
Report Date/Time: Tuesday, June 28, 2016 19:56:57
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.3983**	18.002	173.1	mg/L	0	Standard
Mg	24	58.3	27.6	**0.0205**	0.023	110.8	mg/L	32	Standard
K	39	16.7	45.8	**-0.0851**	0.027	31.2	mg/L	30	Standard
Ca	43	23.3	32.7	**-1.8206**	1.160	63.7	mg/L	48	Standard
Fe	54	118.7	10.7	**-0.0303**	0.010	31.4	mg/L	159	Standard
Fe	57	200.0	10.0	**0.0482**	0.079	163.5	mg/L	195	Standard
Sc-1	45	39738.7	2.1				mg/L	39543	Standard
Cl	35	36016.8	1.9				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	3627.1	7.0				ug/L	1647	Standard
P	31	18584.7	0.6				ug/L	15231	Standard
S	34	2266.8	5.1				ug/L	1845	Standard
Sr	88	53.3	23.6				ug/L	70	Standard
C	12	126.7	32.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	0	Standard
Dy	164	9.0	105.3				mg/L	12	Standard
Ho-1	165	11.7	49.5				mg/L	12	Standard
Er	166	20.0	50.0				mg/L	20	Standard
I	127	1253.4	17.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		90.663	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.501	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW A9 WG572242-03
Report Date/Time: Tuesday, June 28, 2016 19:56:57
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.669
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.956
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW A9 WG572242-03
Report Date/Time: Tuesday, June 28, 2016 19:56:57
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606083901 WG574242-01
Sample Date/Time: Tuesday, June 28, 2016 19:57:52
Number of Replicates: 3
Autosampler Position: 303
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73849.6	1.9				ug/L	81760	Standard
	Be	9	25.0	0.0	**0.0181**	0.001	2.9	ug/L	17	Standard
	Al	27	15521751.4	1.0	**111.4854**	1.467	1.3	ug/L	1410	Standard
	Sc	45	39103.8	3.8				ug/L	39543	Standard
	Ti	47	591.7	5.1	**2.3246**	0.153	6.6	ug/L	39	Standard
	V	51	4219.1	3.4	**0.4282**	0.015	3.5	ug/L	1415	Standard
	Cr	52	8895.6	2.0	**0.2798**	0.040	14.2	ug/L	6946	Standard
	Cr	53	3360.4	10.4	**3.6642**	0.437	11.9	ug/L	328	Standard
	Mn	55	16077.2	1.4	**2.4454**	0.007	0.3	ug/L	823	Standard
	Co	59	1032.0	3.9	**0.1223**	0.006	4.9	ug/L	252	Standard
	Ni	60	3143.3	2.0	**1.6407**	0.018	1.1	ug/L	192	Standard
	Cu	65	3355.0	2.0	**1.7559**	0.016	0.9	ug/L	141	Standard
	Zn	66	14991.7	0.7	**13.4797**	0.137	1.0	ug/L	491	Standard
>	**Ge**	72	492452.2	1.1				ug/L	515087	Standard
	As	75	562.8	15.9	**0.5640**	0.084	14.9	ug/L	20	Standard
	Se	82	67.0	15.0	**0.4895**	0.090	18.3	ug/L	15	Standard
	Se-1	77	207.3	11.8	**2.1604**	0.344	15.9	ug/L	48	Standard
>	**Ga**	71	103.3	23.9				mg/L	90	Standard
	Rb	85	20340.3	1.9				ug/L	23	Standard
	Y	89	439150.3	1.0				ug/L	448510	Standard
>	**Rh**	103	18.3	15.7				ug/L	15	Standard
	Mo	98	1185.2	3.1	**0.4267**	0.016	3.7	ug/L	11	Standard
	Ag	107	124.0	3.2	**0.0026**	0.001	26.5	ug/L	134	Standard
	Cd	111	35.9	29.2	**0.0141**	0.005	34.6	mg/L	7	Standard
	Cd	114	96.0	6.6	**0.0139**	0.001	9.0	ug/L	16	Standard
>	**In**	115	601194.2	0.6				ug/L	611117	Standard
	Sn	118	770.0	3.2	**0.0304**	0.004	13.3	ug/L	518	Standard
	Sb	123	1910.7	20.7	**0.3992**	0.088	22.0	ug/L	127	Standard
	Ba	135	33674.7	0.4	**16.5556**	0.054	0.3	ug/L	45	Standard
	Ce	140	1643.4	3.5				ug/L	298	Standard
>	**Tb**	159	946228.9	1.7				ug/L	977942	Standard
	Ho	165	41.7	36.7				ug/L	12	Standard
	Tl	203	211.3	8.6	**0.0215**	0.002	11.4	ug/L	31	Standard
	Tl	205	188.3	11.1	**0.0247**	0.003	12.3	ug/L	20	Standard
	Pb	206	2721.9	1.8	**0.4638**	0.014	3.1	ug/L	286	Standard
	Pb	207	2208.8	1.5	**0.4164**	0.010	2.4	ug/L	257	Standard
	Pb	208	9386.5	1.8	**0.4607**	0.014	3.1	ug/L	956	Standard
	U	238	2364.2	1.6	**0.3425**	0.008	2.2	ug/L	17	Standard
>	**Bi**	209	532104.5	1.0				ug/L	557080	Standard

Sample ID: L1606083901 WG574242-01
Report Date/Time: Tuesday, June 28, 2016 20:00:08
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.9240**	18.912	173.1	mg/L	0	Standard
Mg	24	14932.4	1.8	**20.0717**	1.109	5.5	mg/L	32	Standard
K	39	851.7	12.5	**2.7753**	0.419	15.1	mg/L	30	Standard
Ca	43	323.3	7.9	**44.3168**	3.500	7.9	mg/L	48	Standard
Fe	54	180.7	15.7	**0.0282**	0.024	84.5	mg/L	159	Standard
Fe	57	355.0	6.1	**0.5869**	0.046	7.9	mg/L	195	Standard
> **Sc-1**	45	39103.8	3.8				mg/L	39543	Standard
Cl	35	38446.7	1.8				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	10497.0	5.2				ug/L	1647	Standard
P	31	21567.0	1.7				ug/L	15231	Standard
S	34	2493.5	2.8				ug/L	1845	Standard
Sr	88	151.7	7.6				ug/L	70	Standard
C	12	210.0	39.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	67.9	17.5				mg/L	12	Standard
Ho-1	165	41.7	36.7				mg/L	12	Standard
Er	166	43.3	93.3				mg/L	20	Standard
I	127	29340.1	3.5				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		90.325	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		95.606	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606083901 WG574242-01
Report Date/Time: Tuesday, June 28, 2016 20:00:08
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.376
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.517
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606083901DP WG574242-04
Sample Date/Time: Tuesday, June 28, 2016 20:01:03
Number of Replicates: 3
Autosampler Position: 304
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74496.0	1.1				ug/L	81760	Standard
	Be	9	20.0	25.0	**0.0124**	0.006	45.8	ug/L	17	Standard
	Al	27	15458448.5	1.4	**110.0484**	0.587	0.5	ug/L	1410	Standard
	Sc	45	38253.3	3.4				ug/L	39543	Standard
	Ti	47	558.7	5.4	**2.1548**	0.155	7.2	ug/L	39	Standard
	V	51	4067.8	2.6	**0.3973**	0.009	2.2	ug/L	1415	Standard
	Cr	52	9024.7	1.6	**0.2811**	0.031	11.1	ug/L	6946	Standard
	Cr	53	3592.1	3.4	**3.8809**	0.090	2.3	ug/L	328	Standard
	Mn	55	13517.4	2.1	**2.0039**	0.068	3.4	ug/L	823	Standard
	Co	59	1022.7	3.5	**0.1189**	0.006	5.0	ug/L	252	Standard
	Ni	60	3054.6	1.9	**1.5691**	0.041	2.6	ug/L	192	Standard
	Cu	65	3452.1	3.2	**1.7838**	0.034	1.9	ug/L	141	Standard
	Zn	66	14844.3	0.5	**13.1506**	0.169	1.3	ug/L	491	Standard
>	**Ge**	72	499109.4	1.4				ug/L	515087	Standard
	As	75	584.5	8.0	**0.5768**	0.041	7.2	ug/L	20	Standard
	Se	82	58.0	10.0	**0.3998**	0.046	11.4	ug/L	15	Standard
	Se-1	77	200.7	10.2	**2.0328**	0.318	15.6	ug/L	48	Standard
>	**Ga**	71	100.0	13.2				mg/L	90	Standard
	Rb	85	19739.5	1.7				ug/L	23	Standard
	Y	89	436250.9	1.8				ug/L	448510	Standard
>	**Rh**	103	38.3	7.5				ug/L	15	Standard
	Mo	98	1219.4	1.0	**0.4525**	0.003	0.7	ug/L	11	Standard
	Ag	107	99.7	5.7	**-0.0007**	0.001	125.7	ug/L	134	Standard
	Cd	111	32.6	12.4	**0.0130**	0.002	14.9	mg/L	7	Standard
	Cd	114	81.5	18.5	**0.0115**	0.003	24.6	ug/L	16	Standard
>	**In**	115	584572.0	1.4				ug/L	611117	Standard
	Sn	118	751.7	6.1	**0.0310**	0.009	27.5	ug/L	518	Standard
	Sb	123	786.1	12.1	**0.1616**	0.023	14.4	ug/L	127	Standard
	Ba	135	32184.1	0.9	**16.2729**	0.095	0.6	ug/L	45	Standard
	Ce	140	490.0	7.4				ug/L	298	Standard
>	**Tb**	159	929479.8	0.3				ug/L	977942	Standard
	Ho	165	31.7	24.1				ug/L	12	Standard
	Tl	203	152.7	10.6	**0.0145**	0.002	14.7	ug/L	31	Standard
	Tl	205	146.7	17.2	**0.0191**	0.004	18.8	ug/L	20	Standard
	Pb	206	2610.2	1.4	**0.4511**	0.005	1.2	ug/L	286	Standard
	Pb	207	2110.5	2.5	**0.4033**	0.013	3.2	ug/L	257	Standard
	Pb	208	8866.0	0.9	**0.4405**	0.006	1.4	ug/L	956	Standard
	U	238	2335.8	3.1	**0.3444**	0.010	2.9	ug/L	17	Standard
>	**Bi**	209	522741.7	0.4				ug/L	557080	Standard

Sample ID: L1606083901DP WG574242-04
Report Date/Time: Tuesday, June 28, 2016 20:03:20
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**11.1196**	19.251	173.1	mg/L	0	Standard
Mg	24	14720.5	1.4	**20.2196**	0.936	4.6	mg/L	32	Standard
K	39	903.4	10.8	**3.0129**	0.317	10.5	mg/L	30	Standard
Ca	43	308.3	1.9	**43.1033**	1.939	4.5	mg/L	48	Standard
Fe	54	132.3	23.0	**-0.0139**	0.024	174.3	mg/L	159	Standard
Fe	57	371.7	11.8	**0.6755**	0.185	27.4	mg/L	195	Standard
> **Sc-1**	45	38253.3	3.4				mg/L	39543	Standard
Cl	35	39853.0	2.0				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	7995.4	3.8				ug/L	1647	Standard
P	31	20689.1	2.4				ug/L	15231	Standard
S	34	2350.2	5.7				ug/L	1845	Standard
Sr	88	115.0	7.5				ug/L	70	Standard
C	12	266.7	11.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	41.7	57.7				mg/L	12	Standard
Ho-1	165	31.7	24.1				mg/L	12	Standard
Er	166	33.3	75.5				mg/L	20	Standard
I	127	28308.1	3.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.115	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.898	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606083901DP WG574242-04
Report Date/Time: Tuesday, June 28, 2016 20:03:20
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.656
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		93.836
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606083901DP WG574242-04
Report Date/Time: Tuesday, June 28, 2016 20:03:20
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606083901S WG574242-05
Sample Date/Time: Tuesday, June 28, 2016 20:04:14
Number of Replicates: 3
Autosampler Position: 305
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77052.1	1.2				ug/L	81760	Standard
	Be	9	48365.4	1.4	**51.6767**	1.282	2.5	ug/L	17	Standard
	Al	27	15205678.0	0.8	**104.6750**	1.999	1.9	ug/L	1410	Standard
	Sc	45	39767.1	2.3				ug/L	39543	Standard
	Ti	47	566.0	2.6	**2.1182**	0.041	1.9	ug/L	39	Standard
	V	51	345428.5	0.5	**49.6572**	0.625	1.3	ug/L	1415	Standard
	Cr	52	342558.9	0.4	**48.3219**	0.886	1.8	ug/L	6946	Standard
	Cr	53	45610.2	2.6	**51.7999**	1.901	3.7	ug/L	328	Standard
	Mn	55	334844.2	1.7	**51.6671**	0.130	0.3	ug/L	823	Standard
	Co	59	340587.2	2.1	**48.5963**	0.256	0.5	ug/L	252	Standard
	Ni	60	92425.4	0.6	**49.1352**	0.675	1.4	ug/L	192	Standard
	Cu	65	94140.2	0.6	**49.6217**	0.865	1.7	ug/L	141	Standard
	Zn	66	68169.3	0.4	**61.5791**	1.173	1.9	ug/L	491	Standard
>	**Ge**	72	513441.7	1.6				ug/L	515087	Standard
	As	75	56899.2	0.6	**50.8164**	0.583	1.1	ug/L	20	Standard
	Se	82	5653.4	1.3	**49.1454**	0.462	0.9	ug/L	15	Standard
	Se-1	77	3932.8	1.5	**51.6632**	1.250	2.4	ug/L	48	Standard
>	**Ga**	71	93.3	35.7				mg/L	90	Standard
	Rb	85	19300.6	2.7				ug/L	23	Standard
	Y	89	425453.4	2.7				ug/L	448510	Standard
>	**Rh**	103	45.0	19.2				ug/L	15	Standard
	Mo	98	1218.6	0.9	**0.4503**	0.008	1.7	ug/L	11	Standard
	Ag	107	314898.7	0.6	**48.8175**	0.102	0.2	ug/L	134	Standard
	Cd	111	99777.7	0.5	**48.5906**	0.622	1.3	mg/L	7	Standard
	Cd	114	243059.8	0.9	**48.6969**	0.795	1.6	ug/L	16	Standard
>	**In**	115	587073.7	0.8				ug/L	611117	Standard
	Sn	118	733.4	7.3	**0.0272**	0.010	36.7	ug/L	518	Standard
	Sb	123	223776.0	1.4	**49.3561**	0.290	0.6	ug/L	127	Standard
	Ba	135	127336.9	0.7	**64.1810**	0.344	0.5	ug/L	45	Standard
	Ce	140	335.0	9.1				ug/L	298	Standard
>	**Tb**	159	937062.8	1.0				ug/L	977942	Standard
	Ho	165	41.7	79.9				ug/L	12	Standard
	Tl	203	401801.7	0.6	**49.2378**	0.579	1.2	ug/L	31	Standard
	Tl	205	351031.6	0.7	**49.5383**	0.691	1.4	ug/L	20	Standard
	Pb	206	267916.7	0.7	**51.1246**	0.591	1.2	ug/L	286	Standard
	Pb	207	234005.3	0.9	**49.5047**	0.372	0.8	ug/L	257	Standard
	Pb	208	924903.3	0.8	**50.5781**	0.765	1.5	ug/L	956	Standard
	U	238	349197.5	0.5	**50.3229**	0.614	1.2	ug/L	17	Standard
>	**Bi**	209	539520.7	0.7				ug/L	557080	Standard

Sample ID: L1606083901S WG574242-05
Report Date/Time: Tuesday, June 28, 2016 20:06:31
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.8832**	18.842	173.1	mg/L	0	Standard
Mg	24	14542.0	3.3	**19.1893**	0.286	1.5	mg/L	32	Standard
K	39	753.4	10.6	**2.3925**	0.305	12.7	mg/L	30	Standard
Ca	43	295.0	12.2	**39.3051**	6.421	16.3	mg/L	48	Standard
Fe	54	140.6	16.7	**-0.0107**	0.020	183.5	mg/L	159	Standard
Fe	57	348.3	9.1	**0.5437**	0.082	15.1	mg/L	195	Standard
> **Sc-1**	45	39767.1	2.3				mg/L	39543	Standard
Cl	35	39860.4	1.2				ug/L	27642	Standard
Kr	83	2.3	49.5				ug/L	2	Standard
Br	81	7565.2	10.7				ug/L	1647	Standard
P	31	21044.6	3.8				ug/L	15231	Standard
S	34	2540.2	5.7				ug/L	1845	Standard
Sr	88	145.0	22.6				ug/L	70	Standard
C	12	196.7	11.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	48.4	54.7				mg/L	12	Standard
Ho-1	165	41.7	79.9				mg/L	12	Standard
Er	166	33.3	45.8				mg/L	20	Standard
I	127	26943.9	1.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.242	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.681	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606083901S WG574242-05
Report Date/Time: Tuesday, June 28, 2016 20:06:31
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.066
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.848
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606083901S WG574242-05
Report Date/Time: Tuesday, June 28, 2016 20:06:31
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606083901SD WG574242-06
Sample Date/Time: Tuesday, June 28, 2016 20:07:26
Number of Replicates: 3
Autosampler Position: 306
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	77184.4	1.1				ug/L	81760	Standard
	Be	9	48315.3	1.2	**51.5258**	0.436	0.8	ug/L	17	Standard
	Al	27	15387465.4	1.3	**105.7299**	0.773	0.7	ug/L	1410	Standard
	Sc	45	40039.5	2.3				ug/L	39543	Standard
	Ti	47	554.0	4.1	**2.1409**	0.093	4.3	ug/L	39	Standard
	V	51	332776.2	1.5	**49.3498**	0.268	0.5	ug/L	1415	Standard
	Cr	52	340776.8	0.8	**49.6175**	0.095	0.2	ug/L	6946	Standard
	Cr	53	45466.5	0.7	**53.2763**	0.775	1.5	ug/L	328	Standard
	Mn	55	324192.4	1.4	**51.6123**	0.258	0.5	ug/L	823	Standard
	Co	59	327045.3	2.1	**48.1461**	0.569	1.2	ug/L	252	Standard
	Ni	60	91364.7	1.1	**50.1099**	0.072	0.1	ug/L	192	Standard
	Cu	65	91854.3	0.9	**49.9479**	0.081	0.2	ug/L	141	Standard
	Zn	66	66866.6	0.8	**62.3204**	0.118	0.2	ug/L	491	Standard
>	Ge	72	497622.9	1.0				ug/L	515087	Standard
	As	75	54845.8	1.7	**50.5323**	0.480	1.0	ug/L	20	Standard
	Se	82	5429.5	1.2	**48.6954**	0.265	0.5	ug/L	15	Standard
	Se-1	77	3850.8	1.9	**52.1872**	0.544	1.0	ug/L	48	Standard
>	Ga	71	85.0	11.8				mg/L	90	Standard
	Rb	85	19015.3	1.2				ug/L	23	Standard
	Y	89	421533.6	0.8				ug/L	448510	Standard
>	Rh	103	55.0	15.7				ug/L	15	Standard
	Mo	98	1205.1	2.4	**0.4408**	0.012	2.7	ug/L	11	Standard
	Ag	107	307002.5	0.8	**47.1551**	0.386	0.8	ug/L	134	Standard
	Cd	111	97425.8	0.9	**47.0062**	0.430	0.9	mg/L	7	Standard
	Cd	114	239795.9	0.5	**47.5975**	0.300	0.6	ug/L	16	Standard
>	In	115	592518.8	0.1				ug/L	611117	Standard
	Sn	118	761.7	9.1	**0.0309**	0.012	38.7	ug/L	518	Standard
	Sb	123	220298.1	1.2	**48.1440**	0.631	1.3	ug/L	127	Standard
	Ba	135	125301.0	0.6	**62.5732**	0.412	0.7	ug/L	45	Standard
	Ce	140	340.0	1.5				ug/L	298	Standard
>	Tb	159	948091.3	1.9				ug/L	977942	Standard
	Ho	165	23.3	65.5				ug/L	12	Standard
	Tl	203	394026.8	0.4	**48.4831**	0.151	0.3	ug/L	31	Standard
	Tl	205	348699.1	1.4	**49.4146**	1.048	2.1	ug/L	20	Standard
	Pb	206	261711.9	1.4	**50.1430**	0.483	1.0	ug/L	286	Standard
	Pb	207	228078.7	1.5	**48.4479**	0.479	1.0	ug/L	257	Standard
	Pb	208	906176.6	0.5	**49.7571**	0.517	1.0	ug/L	956	Standard
	U	238	340376.9	0.5	**49.2541**	0.514	1.0	ug/L	17	Standard
>	Bi	209	537295.6	0.7				ug/L	557080	Standard

Sample ID: L1606083901SD WG574242-06
Report Date/Time: Tuesday, June 28, 2016 20:09:43
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**20.8666**	18.067	86.6	mg/L	0	Standard
Mg	24	14582.0	2.3	**19.1163**	0.419	2.2	mg/L	32	Standard
K	39	795.0	3.8	**2.5139**	0.164	6.5	mg/L	30	Standard
Ca	43	345.0	2.9	**46.4459**	2.697	5.8	mg/L	48	Standard
Fe	54	142.5	14.7	**-0.0097**	0.018	187.3	mg/L	159	Standard
Fe	57	408.3	2.5	**0.7378**	0.065	8.9	mg/L	195	Standard
> **Sc-1**	45	40039.5	2.3				mg/L	39543	Standard
Cl	35	39300.2	1.1				ug/L	27642	Standard
Kr	83	1.7	173.2				ug/L	2	Standard
Br	81	7651.9	3.5				ug/L	1647	Standard
P	31	20634.0	3.2				ug/L	15231	Standard
S	34	2398.5	2.0				ug/L	1845	Standard
Sr	88	180.0	15.5				ug/L	70	Standard
C	12	226.7	27.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	27.4	58.7				mg/L	12	Standard
Ho-1	165	23.3	65.5				mg/L	12	Standard
Er	166	53.3	108.3				mg/L	20	Standard
I	127	27065.8	3.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.404	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.610	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606083901SD WG574242-06
Report Date/Time: Tuesday, June 28, 2016 20:09:43
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.957
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.448
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606083901SD WG574242-06
Report Date/Time: Tuesday, June 28, 2016 20:09:43
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606084301
Sample Date/Time: Tuesday, June 28, 2016 20:10:37
Number of Replicates: 3
Autosampler Position: 307
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80372.6	1.6				ug/L	81760	Standard
	Be	9	41.7	45.4	**0.0330**	0.020	59.2	ug/L	17	Standard
	Al	27	7042054.5	1.4	**46.4726**	1.334	2.9	ug/L	1410	Standard
	Sc	45	38211.6	5.6				ug/L	39543	Standard
	Ti	47	149.3	19.8	**0.4676**	0.133	28.4	ug/L	39	Standard
	V	51	1769.8	3.9	**0.0612**	0.007	10.9	ug/L	1415	Standard
	Cr	52	8360.6	4.6	**0.2060**	0.044	21.1	ug/L	6946	Standard
	Cr	53	938.4	5.4	**0.7826**	0.048	6.1	ug/L	328	Standard
	Mn	55	4412856.2	0.9	**715.8990**	5.361	0.7	ug/L	823	Standard
	Co	59	49628.8	1.9	**7.3983**	0.034	0.5	ug/L	252	Standard
	Ni	60	31518.1	1.1	**17.4994**	0.073	0.4	ug/L	192	Standard
	Cu	65	1083.7	2.3	**0.5092**	0.018	3.5	ug/L	141	Standard
	Zn	66	58737.0	1.3	**55.5464**	0.787	1.4	ug/L	491	Standard
>	**Ge**	72	489683.3	1.5				ug/L	515087	Standard
	As	75	118.1	28.8	**0.1508**	0.032	21.2	ug/L	20	Standard
	Se	82	133.7	3.3	**1.1023**	0.049	4.4	ug/L	15	Standard
	Se-1	77	146.3	10.3	**1.3266**	0.239	18.0	ug/L	48	Standard
>	**Ga**	71	230.0	39.9				mg/L	90	Standard
	Rb	85	13262.5	1.5				ug/L	23	Standard
	Y	89	437622.3	1.3				ug/L	448510	Standard
>	**Rh**	103	40.0	43.3				ug/L	15	Standard
	Mo	98	84.3	2.5	**0.0124**	0.001	5.4	ug/L	11	Standard
	Ag	107	135.7	14.4	**0.0042**	0.003	69.6	ug/L	134	Standard
	Cd	111	325.5	6.3	**0.1508**	0.009	6.1	mg/L	7	Standard
	Cd	114	800.8	5.2	**0.1507**	0.008	5.1	ug/L	16	Standard
>	**In**	115	605376.8	0.3				ug/L	611117	Standard
	Sn	118	716.7	15.8	**0.0204**	0.019	93.4	ug/L	518	Standard
	Sb	123	1846.5	24.8	**0.3824**	0.099	25.9	ug/L	127	Standard
	Ba	135	61196.0	0.7	**29.8977**	0.107	0.4	ug/L	45	Standard
	Ce	140	4327.3	3.4				ug/L	298	Standard
>	**Tb**	159	946751.9	1.1				ug/L	977942	Standard
	Ho	165	113.3	17.8				ug/L	12	Standard
	Tl	203	274.7	6.5	**0.0286**	0.002	7.3	ug/L	31	Standard
	Tl	205	308.3	53.1	**0.0408**	0.023	55.8	ug/L	20	Standard
	Pb	206	584.3	5.1	**0.0470**	0.006	11.8	ug/L	286	Standard
	Pb	207	484.3	1.7	**0.0436**	0.002	3.9	ug/L	257	Standard
	Pb	208	2242.1	13.2	**0.0611**	0.016	25.8	ug/L	956	Standard
	U	238	167.0	68.8	**0.0208**	0.016	78.4	ug/L	17	Standard
>	**Bi**	209	544741.6	0.3				ug/L	557080	Standard

Sample ID: L1606084301
Report Date/Time: Tuesday, June 28, 2016 20:12:54
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	6142.9	1.9	**8.4190**	0.399	4.7	mg/L	32	Standard
K	39	320.0	19.2	**0.9720**	0.149	15.3	mg/L	30	Standard
Ca	43	193.3	18.7	**25.2140**	6.755	26.8	mg/L	48	Standard
Fe	54	272.9	8.2	**0.1197**	0.034	28.2	mg/L	159	Standard
Fe	57	298.3	20.0	**0.4171**	0.202	48.3	mg/L	195	Standard
Sc-1	45	38211.6	5.6				mg/L	39543	Standard
Cl	35	43040.4	2.1				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	3493.7	8.6				ug/L	1647	Standard
P	31	19535.9	3.1				ug/L	15231	Standard
S	34	2298.5	3.2				ug/L	1845	Standard
Sr	88	191.7	4.0				ug/L	70	Standard
C	12	103.3	24.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	171.7	2.4				mg/L	12	Standard
Ho-1	165	113.3	17.8				mg/L	12	Standard
Er	166	103.3	43.6				mg/L	20	Standard
I	127	15184.3	1.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		98.303	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.068	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606084301
Report Date/Time: Tuesday, June 28, 2016 20:12:54
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.061
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.785
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606084301
Report Date/Time: Tuesday, June 28, 2016 20:12:54
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606084301PS WG574570-01
Sample Date/Time: Tuesday, June 28, 2016 20:13:49
Number of Replicates: 3
Autosampler Position: 308
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84132.9	0.3				ug/L	81760	Standard
	Be	9	53800.4	2.5	**52.6389**	1.459	2.8	ug/L	17	Standard
	Al	27	7286262.3	0.5	**45.9217**	0.095	0.2	ug/L	1410	Standard
	Sc	45	39187.3	1.8				ug/L	39543	Standard
	Ti	47	167.0	8.4	**0.5053**	0.064	12.8	ug/L	39	Standard
	V	51	355576.0	1.4	**50.8299**	0.802	1.6	ug/L	1415	Standard
	Cr	52	356673.6	1.0	**50.0684**	1.233	2.5	ug/L	6946	Standard
	Cr	53	44503.6	1.2	**50.2498**	1.575	3.1	ug/L	328	Standard
	Mn	55	4859292.5	1.8	**747.6759**	17.580	2.4	ug/L	823	Standard
	Co	59	417615.7	1.3	**59.2660**	0.892	1.5	ug/L	252	Standard
	Ni	60	125871.4	1.4	**66.5870**	2.033	3.1	ug/L	192	Standard
	Cu	65	97106.5	1.0	**50.8947**	0.717	1.4	ug/L	141	Standard
	Zn	66	127581.2	0.7	**115.3025**	1.913	1.7	ug/L	491	Standard
>	**Ge**	72	516403.0	2.2				ug/L	515087	Standard
	As	75	57691.6	2.0	**51.2302**	0.986	1.9	ug/L	20	Standard
	Se	82	6053.8	2.8	**52.3316**	1.116	2.1	ug/L	15	Standard
	Se-1	77	3949.2	3.8	**51.5591**	1.119	2.2	ug/L	48	Standard
>	**Ga**	71	251.7	10.0				mg/L	90	Standard
	Rb	85	13841.3	3.1				ug/L	23	Standard
	Y	89	444794.9	3.0				ug/L	448510	Standard
>	**Rh**	103	78.3	24.2				ug/L	15	Standard
	Mo	98	86.4	4.3	**0.0128**	0.001	7.7	ug/L	11	Standard
	Ag	107	327259.4	2.3	**48.6313**	0.643	1.3	ug/L	134	Standard
	Cd	111	105349.6	1.3	**49.1793**	0.322	0.7	mg/L	7	Standard
	Cd	114	257271.1	1.0	**49.4173**	1.101	2.2	ug/L	16	Standard
>	**In**	115	612449.2	2.0				ug/L	611117	Standard
	Sn	118	941.7	3.9	**0.0570**	0.004	7.5	ug/L	518	Standard
	Sb	123	235069.6	1.3	**49.7050**	0.359	0.7	ug/L	127	Standard
	Ba	135	161643.1	1.4	**78.1070**	0.551	0.7	ug/L	45	Standard
	Ce	140	4429.0	1.8				ug/L	298	Standard
>	**Tb**	159	958659.2	2.3				ug/L	977942	Standard
	Ho	165	101.7	15.0				ug/L	12	Standard
	Tl	203	432017.4	2.0	**50.9163**	0.725	1.4	ug/L	31	Standard
	Tl	205	384793.0	2.1	**52.2411**	1.764	3.4	ug/L	20	Standard
	Pb	206	286141.9	1.8	**52.5195**	0.925	1.8	ug/L	286	Standard
	Pb	207	249240.4	1.8	**50.7143**	0.544	1.1	ug/L	257	Standard
	Pb	208	995726.4	2.1	**52.3723**	1.044	2.0	ug/L	956	Standard
	U	238	370701.4	1.5	**51.3906**	1.394	2.7	ug/L	17	Standard
>	**Bi**	209	561025.3	2.5				ug/L	557080	Standard

Sample ID: L1606084301PS WG574570-01
Report Date/Time: Tuesday, June 28, 2016 20:16:06
Page 1

Approved: June 29, 2016

Na	23	5.0	0.0	**32.4034**	0.594	1.8	mg/L	0	Standard
Mg	24	6414.7	1.1	**8.5623**	0.246	2.9	mg/L	32	Standard
K	39	321.7	9.1	**0.9561**	0.110	11.5	mg/L	30	Standard
Ca	43	185.0	8.1	**22.9936**	1.940	8.4	mg/L	48	Standard
Fe	54	250.2	14.1	**0.0912**	0.029	32.3	mg/L	159	Standard
Fe	57	306.7	15.8	**0.4183**	0.145	34.7	mg/L	195	Standard
> **Sc-1**	45	39187.3	1.8				mg/L	39543	Standard
Cl	35	43079.8	1.6				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	3410.4	11.2				ug/L	1647	Standard
P	31	19879.7	4.3				ug/L	15231	Standard
S	34	2313.5	2.6				ug/L	1845	Standard
Sr	88	201.7	5.2				ug/L	70	Standard
C	12	140.0	32.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0					mg/L	0	Standard
Dy	164	196.2	20.7				mg/L	12	Standard
Ho-1	165	101.7	15.0				mg/L	12	Standard
Er	166	80.0	76.0				mg/L	20	Standard
I	127	16267.1	4.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.902	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.256	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606084301PS WG574570-01
Report Date/Time: Tuesday, June 28, 2016 20:16:06
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	100.218
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	100.708
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606084301SDL WG574570-02
Sample Date/Time: Tuesday, June 28, 2016 20:17:00
Number of Replicates: 3
Autosampler Position: 309
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84610.6	1.2				ug/L	81760	Standard
	Be	9	31.7	24.1	**0.0210**	0.007	33.6	ug/L	17	Standard
	Al	27	1535529.0	0.5	**9.6123**	0.084	0.9	ug/L	1410	Standard
	Sc	45	40586.0	2.0				ug/L	39543	Standard
	Ti	47	54.0	14.5	**0.0442**	0.029	66.4	ug/L	39	Standard
	V	51	1510.6	4.6	**0.0030**	0.011	356.0	ug/L	1415	Standard
	Cr	52	7700.6	1.7	**0.0097**	0.022	221.3	ug/L	6946	Standard
	Cr	53	548.3	6.2	**0.2609**	0.037	14.0	ug/L	328	Standard
	Mn	55	927997.3	1.4	**137.8897**	1.374	1.0	ug/L	823	Standard
	Co	59	10741.5	3.1	**1.4429**	0.042	2.9	ug/L	252	Standard
	Ni	60	6856.2	1.9	**3.4057**	0.060	1.7	ug/L	192	Standard
	Cu	65	384.3	5.5	**0.1042**	0.011	10.3	ug/L	141	Standard
	Zn	66	13504.0	1.8	**11.0531**	0.162	1.5	ug/L	491	Standard
>	**Ge**	72	534132.0	0.6				ug/L	515087	Standard
	As	75	-8.2	426.8	**0.0333**	0.030	90.7	ug/L	20	Standard
	Se	82	34.3	14.2	**0.1678**	0.041	24.5	ug/L	15	Standard
	Se-1	77	77.7	7.3	**0.2756**	0.079	28.6	ug/L	48	Standard
>	**Ga**	71	133.3	12.1				mg/L	90	Standard
	Rb	85	2885.3	6.5				ug/L	23	Standard
	Y	89	438369.8	1.0				ug/L	448510	Standard
>	**Rh**	103	21.7	13.3				ug/L	15	Standard
	Mo	98	33.0	37.7	**-0.0070**	0.005	64.6	ug/L	11	Standard
	Ag	107	108.0	7.3	**-0.0003**	0.002	523.1	ug/L	134	Standard
	Cd	111	82.3	4.3	**0.0351**	0.001	2.2	mg/L	7	Standard
	Cd	114	247.3	23.3	**0.0421**	0.011	25.9	ug/L	16	Standard
>	**In**	115	618448.1	2.5				ug/L	611117	Standard
	Sn	118	536.7	11.9	**-0.0124**	0.008	66.9	ug/L	518	Standard
	Sb	123	2208.8	28.0	**0.4480**	0.119	26.5	ug/L	127	Standard
	Ba	135	12894.8	1.5	**6.1480**	0.082	1.3	ug/L	45	Standard
	Ce	140	871.7	9.1				ug/L	298	Standard
>	**Tb**	159	974992.3	2.7				ug/L	977942	Standard
	Ho	165	33.3	48.2				ug/L	12	Standard
	Tl	203	98.0	7.7	**0.0064**	0.001	11.0	ug/L	31	Standard
	Tl	205	90.0	24.2	**0.0096**	0.003	27.9	ug/L	20	Standard
	Pb	206	392.7	2.7	**0.0066**	0.003	42.8	ug/L	286	Standard
	Pb	207	318.0	2.7	**0.0050**	0.001	23.7	ug/L	257	Standard
	Pb	208	1352.7	4.0	**0.0088**	0.002	18.5	ug/L	956	Standard
	U	238	56.0	29.8	**0.0046**	0.002	47.0	ug/L	17	Standard
>	**Bi**	209	576774.8	1.6				ug/L	557080	Standard

Sample ID: L1606084301SDL WG574570-02
Report Date/Time: Tuesday, June 28, 2016 20:19:17
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.6259**	18.396	173.1	mg/L	0	Standard
Mg	24	1335.1	0.6	**1.6748**	0.038	2.2	mg/L	32	Standard
K	39	83.3	27.1	**0.1320**	0.069	52.2	mg/L	30	Standard
Ca	43	66.7	4.3	**4.5227**	0.490	10.8	mg/L	48	Standard
Fe	54	158.9	15.8	**0.0030**	0.021	715.4	mg/L	159	Standard
Fe	57	226.7	20.8	**0.1225**	0.167	136.7	mg/L	195	Standard
> **Sc-1**	45	40586.0	2.0				mg/L	39543	Standard
Cl	35	41629.1	0.8				ug/L	27642	Standard
Kr	83	4.7	24.7				ug/L	2	Standard
Br	81	2016.8	5.8				ug/L	1647	Standard
P	31	19454.2	1.7				ug/L	15231	Standard
S	34	2286.8	4.6				ug/L	1845	Standard
Sr	88	108.3	9.6				ug/L	70	Standard
C	12	106.7	39.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	58.6	31.0				mg/L	12	Standard
Ho-1	165	33.3	48.2				mg/L	12	Standard
Er	166	30.0	57.7				mg/L	20	Standard
I	127	4460.7	6.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		103.486	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		103.697	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606084301SDL WG574570-02
Report Date/Time: Tuesday, June 28, 2016 20:19:17
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.200
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.535
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606084301SDL WG574570-02
Sample Date/Time: Tuesday, June 28, 2016 20:20:12
Number of Replicates: 3
Autosampler Position: 310
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84917.5	3.2				ug/L	81760	Standard
	Be	9	15.0	66.7	**0.0050**	0.010	200.3	ug/L	17	Standard
	Al	27	312447.8	1.9	**1.9378**	0.034	1.7	ug/L	1410	Standard
	Sc	45	41284.5	3.3				ug/L	39543	Standard
	Ti	47	45.7	48.8	**0.0117**	0.082	700.7	ug/L	39	Standard
	V	51	1463.2	5.6	**-0.0026**	0.007	263.9	ug/L	1415	Standard
	Cr	52	7659.3	1.8	**0.0098**	0.004	44.8	ug/L	6946	Standard
	Cr	53	423.3	16.3	**0.1274**	0.086	67.2	ug/L	328	Standard
	Mn	55	187375.9	0.8	**27.8813**	0.482	1.7	ug/L	823	Standard
	Co	59	2308.5	0.8	**0.2873**	0.008	2.8	ug/L	252	Standard
	Ni	60	1540.7	3.0	**0.6883**	0.022	3.2	ug/L	192	Standard
	Cu	65	216.7	7.3	**0.0197**	0.006	31.0	ug/L	141	Standard
	Zn	66	4319.6	1.9	**2.9936**	0.025	0.8	ug/L	491	Standard
>	**Ge**	72	531260.9	2.2				ug/L	515087	Standard
	As	75	-47.8	76.7	**-0.0014**	0.032	2343.8	ug/L	20	Standard
	Se	82	17.0	17.8	**0.0237**	0.023	96.9	ug/L	15	Standard
	Se-1	77	59.0	6.1	**0.0400**	0.030	74.0	ug/L	48	Standard
>	**Ga**	71	86.7	31.8				mg/L	90	Standard
	Rb	85	596.7	10.5				ug/L	23	Standard
	Y	89	453907.4	1.0				ug/L	448510	Standard
>	**Rh**	103	8.3	34.6				ug/L	15	Standard
	Mo	98	28.3	9.0	**-0.0090**	0.001	10.4	ug/L	11	Standard
	Ag	107	92.0	2.2	**-0.0030**	0.000	8.2	ug/L	134	Standard
	Cd	111	15.3	15.1	**0.0040**	0.001	27.2	mg/L	7	Standard
	Cd	114	38.2	13.4	**0.0022**	0.001	41.3	ug/L	16	Standard
>	**In**	115	632674.4	0.8				ug/L	611117	Standard
	Sn	118	625.0	9.7	**0.0002**	0.010	5977.1	ug/L	518	Standard
	Sb	123	614.9	31.7	**0.1128**	0.039	34.5	ug/L	127	Standard
	Ba	135	2594.2	0.9	**1.1887**	0.008	0.6	ug/L	45	Standard
	Ce	140	198.3	20.2				ug/L	298	Standard
>	**Tb**	159	988615.6	0.9				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	26.7	16.9	**-0.0018**	0.000	28.3	ug/L	31	Standard
	Tl	205	16.7	75.5	**-0.0001**	0.002	1448.6	ug/L	20	Standard
	Pb	206	345.0	3.1	**-0.0028**	0.002	58.3	ug/L	286	Standard
	Pb	207	303.3	7.6	**0.0012**	0.004	352.8	ug/L	257	Standard
	Pb	208	1152.7	2.7	**-0.0023**	0.002	71.0	ug/L	956	Standard
	U	238	16.0	21.7	**-0.0008**	0.000	54.8	ug/L	17	Standard
>	**Bi**	209	585233.9	0.6				ug/L	557080	Standard

Sample ID: L1606084301SDL WG574570-02
Report Date/Time: Tuesday, June 28, 2016 20:22:29
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	291.7	12.4	**0.3140**	0.035	11.1	mg/L	32	Standard
K	39	48.3	6.0	**0.0153**	0.013	86.9	mg/L	30	Standard
Ca	43	35.0	28.6	**-0.2811**	1.306	464.6	mg/L	48	Standard
Fe	54	153.8	34.6	**-0.0038**	0.045	1180.2	mg/L	159	Standard
Fe	57	170.0	8.8	**-0.0754**	0.036	48.4	mg/L	195	Standard
> **Sc-1**	45	41284.5	3.3				mg/L	39543	Standard
Cl	35	41099.7	0.8				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	1820.1	5.7				ug/L	1647	Standard
P	31	19375.7	1.6				ug/L	15231	Standard
S	34	2545.2	4.1				ug/L	1845	Standard
Sr	88	58.3	27.6				ug/L	70	Standard
C	12	110.0	39.6				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	28.2	93.4				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	36.7	15.7				mg/L	20	Standard
I	127	1790.1	13.3				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		103.862	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		103.140	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606084301SDL WG574570-02
Report Date/Time: Tuesday, June 28, 2016 20:22:29
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.528
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.054
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 20:23:25
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	82271.2	2.6				ug/L	81760	Standard
	Be	9	51878.7	2.6	**51.9060**	0.474	0.9	ug/L	17	Standard
	Al	27	7912369.8	3.4	**50.9914**	0.481	0.9	ug/L	1410	Standard
	Sc	45	40149.8	1.3				ug/L	39543	Standard
	Ti	47	25445.3	0.9	**101.2486**	0.646	0.6	ug/L	39	Standard
	V	51	355257.7	1.2	**50.4501**	0.253	0.5	ug/L	1415	Standard
	Cr	52	364176.4	2.0	**50.7897**	0.304	0.6	ug/L	6946	Standard
	Cr	53	45239.1	0.4	**50.7395**	0.592	1.2	ug/L	328	Standard
	Mn	55	333791.7	0.4	**50.8845**	0.584	1.1	ug/L	823	Standard
	Co	59	367088.8	1.7	**51.7480**	0.425	0.8	ug/L	252	Standard
	Ni	60	95399.1	1.8	**50.0959**	0.391	0.8	ug/L	192	Standard
	Cu	65	97309.2	2.1	**50.6606**	0.365	0.7	ug/L	141	Standard
	Zn	66	56912.4	2.6	**50.6269**	0.602	1.2	ug/L	491	Standard
>	Ge	72	519736.7	1.5				ug/L	515087	Standard
	As	75	56554.6	1.5	**49.8917**	0.034	0.1	ug/L	20	Standard
	Se	82	5843.7	2.8	**50.1784**	0.757	1.5	ug/L	15	Standard
	Se-1	77	3965.5	1.4	**51.4490**	0.315	0.6	ug/L	48	Standard
>	Ga	71	118.3	13.6				mg/L	90	Standard
	Rb	85	848.4	6.2				ug/L	23	Standard
	Y	89	451425.6	1.6				ug/L	448510	Standard
>	Rh	103	31.7	71.2				ug/L	15	Standard
	Mo	98	272060.7	0.6	**99.5752**	0.551	0.6	ug/L	11	Standard
	Ag	107	339081.9	1.7	**49.9711**	1.119	2.2	ug/L	134	Standard
	Cd	111	106686.1	0.8	**49.3843**	0.588	1.2	mg/L	7	Standard
	Cd	114	259655.5	1.7	**49.4494**	1.121	2.3	ug/L	16	Standard
>	In	115	617617.8	0.6				ug/L	611117	Standard
	Sn	118	302897.0	2.2	**50.5905**	1.415	2.8	ug/L	518	Standard
	Sb	123	239436.4	1.3	**50.2037**	0.916	1.8	ug/L	127	Standard
	Ba	135	101930.1	0.6	**48.8299**	0.530	1.1	ug/L	45	Standard
	Ce	140	63.3	9.1				ug/L	298	Standard
>	Tb	159	978625.5	0.9				ug/L	977942	Standard
	Ho	165	8.3	69.3				ug/L	12	Standard
	Tl	203	434700.2	0.8	**50.4033**	0.455	0.9	ug/L	31	Standard
	Tl	205	381183.1	1.9	**50.8972**	0.874	1.7	ug/L	20	Standard
	Pb	206	277544.1	1.7	**50.1113**	0.892	1.8	ug/L	286	Standard
	Pb	207	250965.9	2.0	**50.2393**	1.021	2.0	ug/L	257	Standard
	Pb	208	994830.3	0.3	**51.4752**	0.269	0.5	ug/L	956	Standard
	U	238	372750.7	1.1	**50.8268**	0.592	1.2	ug/L	17	Standard
>	Bi	209	570174.9	0.2				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 20:25:42
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4207.3	2.1	**5.4598**	0.160	2.9	mg/L	32	Standard
K	39	1353.4	1.9	**4.3620**	0.034	0.8	mg/L	30	Standard
Ca	43	68.3	23.5	**4.8558**	2.284	47.0	mg/L	48	Standard
Fe	54	6096.4	4.5	**5.2968**	0.235	4.4	mg/L	159	Standard
Fe	57	1726.8	6.1	**5.1204**	0.425	8.3	mg/L	195	Standard
Sc-1	45	40149.8	1.3				mg/L	39543	Standard
Cl	35	38223.5	2.9				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	1710.1	6.5				ug/L	1647	Standard
P	31	20984.5	3.6				ug/L	15231	Standard
S	34	2836.9	3.0				ug/L	1845	Standard
Sr	88	73.3	32.2				ug/L	70	Standard
C	12	176.7	8.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	26.3	80.0				mg/L	12	Standard
Ho-1	165	8.3	69.3				mg/L	12	Standard
Er	166	6.7	86.6				mg/L	20	Standard
I	127	1388.4	24.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	103.812		
Al	27	101.983		
Sc	45			
Ti	47	101.249		
V	51	100.900		
Cr	52	101.579		
Cr	53			
Mn	55	101.769		
Co	59	103.496		
Ni	60	100.192		
Cu	65	101.321		
Zn	66	101.254		
Ge	72		100.903	
As	75	99.783		
Se	82	100.357		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 20:25:42
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	99.575		
Ag	107	99.942		
Cd	111	98.769		
Cd	114			
In	115			101.064
Sn	118	101.181		
Sb	123	100.407		
Ba	135	97.660		
Ce	140			
Tb	159			
Ho	165			
Tl	203	100.807		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.950		
U	238	101.654		
Bi	209			102.351
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 20:25:42
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 20:26:37
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77291.7	2.7				ug/L	81760	Standard
	Be	9	25.0	40.0	**0.0170**	0.011	63.1	ug/L	17	Standard
	Al	27	2576.9	24.9	**0.0032**	0.004	126.1	ug/L	1410	Standard
	Sc	45	39050.3	1.7				ug/L	39543	Standard
	Ti	47	33.7	13.4	**-0.0246**	0.019	78.0	ug/L	39	Standard
	V	51	1167.6	8.0	**-0.0323**	0.013	41.5	ug/L	1415	Standard
	Cr	52	6011.5	1.2	**-0.1610**	0.012	7.3	ug/L	6946	Standard
	Cr	53	398.3	5.1	**0.1293**	0.022	17.0	ug/L	328	Standard
	Mn	55	883.7	9.4	**-0.0071**	0.013	183.2	ug/L	823	Standard
	Co	59	243.0	13.8	**0.0045**	0.005	107.7	ug/L	252	Standard
	Ni	60	165.0	10.4	**-0.0141**	0.009	65.1	ug/L	192	Standard
	Cu	65	129.3	4.3	**-0.0202**	0.003	14.1	ug/L	141	Standard
	Zn	66	451.0	1.9	**-0.4007**	0.009	2.3	ug/L	491	Standard
>	**Ge**	72	496420.0	0.5				ug/L	515087	Standard
	As	75	-52.8	22.7	**-0.0085**	0.011	133.2	ug/L	20	Standard
	Se	82	7.8	31.0	**-0.0491**	0.022	44.9	ug/L	15	Standard
	Se-1	77	52.0	3.8	**-0.0027**	0.026	948.5	ug/L	48	Standard
>	**Ga**	71	61.7	18.7				mg/L	90	Standard
	Rb	85	18.3	103.3				ug/L	23	Standard
	Y	89	442161.2	1.2				ug/L	448510	Standard
>	**Rh**	103	6.7	43.3				ug/L	15	Standard
	Mo	98	281.8	35.5	**0.0839**	0.034	41.1	ug/L	11	Standard
	Ag	107	122.3	22.2	**0.0018**	0.004	200.6	ug/L	134	Standard
	Cd	111	14.5	44.1	**0.0038**	0.003	74.5	mg/L	7	Standard
	Cd	114	43.1	30.3	**0.0033**	0.002	70.6	ug/L	16	Standard
>	**In**	115	615841.7	2.0				ug/L	611117	Standard
	Sn	118	753.4	17.4	**0.0248**	0.024	98.6	ug/L	518	Standard
	Sb	123	596.8	25.3	**0.1131**	0.034	30.2	ug/L	127	Standard
	Ba	135	38.0	16.0	**-0.0069**	0.003	39.0	ug/L	45	Standard
	Ce	140	8.3	91.7				ug/L	298	Standard
>	**Tb**	159	956084.4	3.2				ug/L	977942	Standard
	Ho	165	5.0	100.0				ug/L	12	Standard
	Tl	203	73.0	30.8	**0.0039**	0.003	64.7	ug/L	31	Standard
	Tl	205	60.0	44.1	**0.0059**	0.003	58.6	ug/L	20	Standard
	Pb	206	307.3	5.4	**-0.0063**	0.002	32.6	ug/L	286	Standard
	Pb	207	254.3	1.8	**-0.0055**	0.000	8.7	ug/L	257	Standard
	Pb	208	1093.3	8.1	**-0.0021**	0.004	180.6	ug/L	956	Standard
	U	238	62.0	42.0	**0.0057**	0.004	62.4	ug/L	17	Standard
>	**Bi**	209	552726.9	1.8				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 20:28:53
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.8094**	18.714	173.1	mg/L	0	Standard
Mg	24	20.0	100.0	**-0.0300**	0.027	91.7	mg/L	32	Standard
K	39	31.7	50.8	**-0.0336**	0.053	157.9	mg/L	30	Standard
Ca	43	16.7	17.3	**-2.7773**	0.490	17.6	mg/L	48	Standard
Fe	54	136.0	26.0	**-0.0122**	0.034	282.8	mg/L	159	Standard
Fe	57	233.3	19.3	**0.1724**	0.147	85.5	mg/L	195	Standard
Sc-1	45	39050.3	1.7				mg/L	39543	Standard
Cl	35	37106.7	0.9				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	1596.8	5.3				ug/L	1647	Standard
P	31	19597.7	1.6				ug/L	15231	Standard
S	34	2643.6	3.6				ug/L	1845	Standard
Sr	88	53.3	61.0				ug/L	70	Standard
C	12	110.0	32.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	0	Standard
Dy	164	9.4	104.3				mg/L	12	Standard
Ho-1	165	5.0	100.0				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	1173.4	25.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.376	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 20:28:53
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.773
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.219
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606079401
Sample Date/Time: Tuesday, June 28, 2016 20:29:50
Number of Replicates: 3
Autosampler Position: 311
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76244.9	4.1				ug/L	81760	Standard
	Be	9	31.7	63.8	**0.0239**	0.020	83.5	ug/L	17	Standard
	Al	27	4859544.3	1.6	**33.8379**	1.720	5.1	ug/L	1410	Standard
	Sc	45	41316.3	3.1				ug/L	39543	Standard
	Ti	47	1613.1	8.6	**6.3222**	0.521	8.2	ug/L	39	Standard
	V	51	5296.6	3.8	**0.5556**	0.022	4.0	ug/L	1415	Standard
	Cr	52	9948.6	1.0	**0.3727**	0.012	3.3	ug/L	6946	Standard
	Cr	53	1300.1	3.5	**1.1398**	0.032	2.8	ug/L	328	Standard
	Mn	55	10830112.5	0.4	**1671.2513**	23.447	1.4	ug/L	823	Standard
	Co	59	98900.4	0.9	**14.0508**	0.072	0.5	ug/L	252	Standard
	Ni	60	5918.2	0.8	**3.0389**	0.039	1.3	ug/L	192	Standard
	Cu	65	1492.7	4.1	**0.6949**	0.022	3.2	ug/L	141	Standard
	Zn	66	3606.1	2.5	**2.4634**	0.043	1.7	ug/L	491	Standard
>	**Ge**	72	514893.7	1.4				ug/L	515087	Standard
	As	75	2838.5	0.7	**2.5661**	0.029	1.1	ug/L	20	Standard
	Se	82	67.5	5.6	**0.4677**	0.041	8.7	ug/L	15	Standard
	Se-1	77	72.7	2.9	**0.2462**	0.031	12.7	ug/L	48	Standard
>	**Ga**	71	795.0	2.5				mg/L	90	Standard
	Rb	85	32504.8	2.3				ug/L	23	Standard
	Y	89	435925.3	2.8				ug/L	448510	Standard
>	**Rh**	103	93.3	3.1				ug/L	15	Standard
	Mo	98	343.6	1.5	**0.1114**	0.005	4.3	ug/L	11	Standard
	Ag	107	117.7	11.3	**0.0018**	0.002	118.8	ug/L	134	Standard
	Cd	111	40.7	19.8	**0.0166**	0.004	21.3	mg/L	7	Standard
	Cd	114	97.5	38.9	**0.0145**	0.008	54.8	ug/L	16	Standard
>	**In**	115	595657.1	2.2				ug/L	611117	Standard
	Sn	118	961.7	5.9	**0.0649**	0.008	12.2	ug/L	518	Standard
	Sb	123	345.9	22.0	**0.0622**	0.015	24.5	ug/L	127	Standard
	Ba	135	208994.2	0.7	**103.8626**	2.069	2.0	ug/L	45	Standard
	Ce	140	33943.0	0.8				ug/L	298	Standard
>	**Tb**	159	940708.8	1.0				ug/L	977942	Standard
	Ho	165	610.0	7.0				ug/L	12	Standard
	Tl	203	714.0	4.8	**0.0811**	0.003	4.2	ug/L	31	Standard
	Tl	205	651.7	8.6	**0.0881**	0.009	9.8	ug/L	20	Standard
	Pb	206	1354.7	2.2	**0.1905**	0.003	1.6	ug/L	286	Standard
	Pb	207	1119.4	1.4	**0.1747**	0.005	2.8	ug/L	257	Standard
	Pb	208	4748.3	2.1	**0.1947**	0.004	1.9	ug/L	956	Standard
	U	238	416.3	2.3	**0.0559**	0.002	3.0	ug/L	17	Standard
>	**Bi**	209	549596.9	1.2				ug/L	557080	Standard

Sample ID: L1606079401
Report Date/Time: Tuesday, June 28, 2016 20:32:07
Page 1

Approved: June 29, 2016

Na	23	6.7	114.6	**41.1239**	47.883	116.4	mg/L	0	Standard
Mg	24	3483.7	2.6	**4.3856**	0.231	5.3	mg/L	32	Standard
K	39	496.7	14.5	**1.4639**	0.222	15.1	mg/L	30	Standard
Ca	43	190.0	16.0	**22.2436**	4.083	18.4	mg/L	48	Standard
Fe	54	7377.0	1.5	**6.2559**	0.190	3.0	mg/L	159	Standard
Fe	57	2103.5	3.2	**6.1801**	0.414	6.7	mg/L	195	Standard
Sc-1	45	41316.3	3.1				mg/L	39543	Standard
Cl	35	35106.6	2.5				ug/L	27642	Standard
Kr	83	2.3	107.9				ug/L	2	Standard
Br	81	13476.0	5.7				ug/L	1647	Standard
P	31	22598.5	5.8				ug/L	15231	Standard
S	34	2571.9	2.2				ug/L	1845	Standard
Sr	88	433.3	9.8				ug/L	70	Standard
C	12	150.0	35.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	882.4	15.5				mg/L	12	Standard
Ho-1	165	610.0	7.0				mg/L	12	Standard
Er	166	576.7	6.6				mg/L	20	Standard
I	127	50022.6	4.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.254	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.963	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606079401
Report Date/Time: Tuesday, June 28, 2016 20:32:07
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.470
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	98.657
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606079401
Report Date/Time: Tuesday, June 28, 2016 20:32:07
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606079402
Sample Date/Time: Tuesday, June 28, 2016 20:33:01
Number of Replicates: 3
Autosampler Position: 312
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75678.5	1.2				ug/L	81760	Standard
	Be	9	111.7	13.7	**0.1118**	0.017	15.0	ug/L	17	Standard
	Al	27	1429595.1	1.1	**10.0072**	0.228	2.3	ug/L	1410	Standard
	Sc	45	39138.8	1.6				ug/L	39543	Standard
	Ti	47	278.3	21.3	**0.9895**	0.212	21.4	ug/L	39	Standard
	V	51	1557.1	6.8	**0.0252**	0.009	34.9	ug/L	1415	Standard
	Cr	52	8352.6	6.5	**0.1848**	0.041	22.0	ug/L	6946	Standard
	Cr	53	1770.1	12.2	**1.7430**	0.191	11.0	ug/L	328	Standard
	Mn	55	34175.5	2.4	**5.3156**	0.224	4.2	ug/L	823	Standard
	Co	59	3266.0	5.1	**0.4499**	0.010	2.1	ug/L	252	Standard
	Ni	60	1953.8	3.9	**0.9695**	0.017	1.8	ug/L	192	Standard
	Cu	65	1358.7	2.9	**0.6499**	0.011	1.6	ug/L	141	Standard
	Zn	66	4870.4	2.4	**3.7758**	0.058	1.5	ug/L	491	Standard
>	**Ge**	72	497312.9	3.3				ug/L	515087	Standard
	As	75	54.7	94.2	**0.0915**	0.049	53.2	ug/L	20	Standard
	Se	82	40.7	27.3	**0.2488**	0.112	45.1	ug/L	15	Standard
	Se-1	77	77.0	9.4	**0.3389**	0.077	22.7	ug/L	48	Standard
>	**Ga**	71	155.0	28.1				mg/L	90	Standard
	Rb	85	30158.3	1.3				ug/L	23	Standard
	Y	89	432228.9	1.1				ug/L	448510	Standard
>	**Rh**	103	11.7	49.5				ug/L	15	Standard
	Mo	98	72.4	21.3	**0.0082**	0.006	70.5	ug/L	11	Standard
	Ag	107	118.3	24.0	**0.0017**	0.004	249.1	ug/L	134	Standard
	Cd	111	44.6	10.6	**0.0183**	0.002	12.1	mg/L	7	Standard
	Cd	114	123.7	31.0	**0.0193**	0.007	38.8	ug/L	16	Standard
>	**In**	115	600554.1	0.2				ug/L	611117	Standard
	Sn	118	733.4	8.7	**0.0243**	0.011	44.6	ug/L	518	Standard
	Sb	123	231.2	28.1	**0.0370**	0.014	37.8	ug/L	127	Standard
	Ba	135	65483.6	1.3	**32.2516**	0.425	1.3	ug/L	45	Standard
	Ce	140	2745.2	3.0				ug/L	298	Standard
>	**Tb**	159	936903.6	1.4				ug/L	977942	Standard
	Ho	165	75.0	17.6				ug/L	12	Standard
	Tl	203	221.3	13.5	**0.0222**	0.003	15.7	ug/L	31	Standard
	Tl	205	186.7	13.2	**0.0239**	0.003	14.3	ug/L	20	Standard
	Pb	206	523.0	7.4	**0.0358**	0.006	17.6	ug/L	286	Standard
	Pb	207	464.7	6.5	**0.0400**	0.008	19.0	ug/L	257	Standard
	Pb	208	1837.7	5.9	**0.0397**	0.005	13.6	ug/L	956	Standard
	U	238	125.7	23.1	**0.0151**	0.004	27.1	ug/L	17	Standard
>	**Bi**	209	541937.8	1.7				ug/L	557080	Standard

Sample ID: L1606079402
Report Date/Time: Tuesday, June 28, 2016 20:35:18
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	2726.9	0.4	**3.6111**	0.076	2.1	mg/L	32	Standard
K	39	506.7	5.1	**1.5885**	0.091	5.7	mg/L	30	Standard
Ca	43	36.7	15.7	**0.2847**	0.917	322.3	mg/L	48	Standard
Fe	54	182.6	18.3	**0.0302**	0.033	110.1	mg/L	159	Standard
Fe	57	203.3	25.6	**0.0706**	0.187	264.3	mg/L	195	Standard
Sc-1	45	39138.8	1.6				mg/L	39543	Standard
Cl	35	34567.4	3.2				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	6558.1	13.2				ug/L	1647	Standard
P	31	21156.4	3.6				ug/L	15231	Standard
S	34	2326.8	9.7				ug/L	1845	Standard
Sr	88	76.7	13.6				ug/L	70	Standard
C	12	236.7	44.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	150.6	9.6				mg/L	12	Standard
Ho-1	165	75.0	17.6				mg/L	12	Standard
Er	166	56.7	40.8				mg/L	20	Standard
I	127	9424.6	4.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		92.562	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.549	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606079402
Report Date/Time: Tuesday, June 28, 2016 20:35:18
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.272
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.282
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: L1606079402
Report Date/Time: Tuesday, June 28, 2016 20:35:18
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606079403
Sample Date/Time: Tuesday, June 28, 2016 20:36:13
Number of Replicates: 3
Autosampler Position: 313
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76753.9	0.3				ug/L	81760	Standard
	Be	9	53.3	32.9	**0.0475**	0.019	39.7	ug/L	17	Standard
	Al	27	2019995.9	2.1	**13.9455**	0.321	2.3	ug/L	1410	Standard
	Sc	45	39648.5	2.7				ug/L	39543	Standard
	Ti	47	884.4	21.1	**3.3748**	0.755	22.4	ug/L	39	Standard
	V	51	4216.0	3.5	**0.3970**	0.022	5.5	ug/L	1415	Standard
	Cr	52	11350.6	1.0	**0.5656**	0.020	3.6	ug/L	6946	Standard
	Cr	53	3082.0	1.2	**3.1525**	0.043	1.4	ug/L	328	Standard
	Mn	55	1462715.3	1.2	**224.3814**	2.702	1.2	ug/L	823	Standard
	Co	59	3343.0	3.0	**0.4421**	0.013	2.9	ug/L	252	Standard
	Ni	60	3343.7	2.8	**1.6618**	0.047	2.8	ug/L	192	Standard
	Cu	65	1050.7	3.0	**0.4594**	0.015	3.2	ug/L	141	Standard
	Zn	66	4988.5	1.4	**3.7017**	0.077	2.1	ug/L	491	Standard
>	**Ge**	72	517603.4	0.3				ug/L	515087	Standard
	As	75	193.9	14.0	**0.2119**	0.024	11.1	ug/L	20	Standard
	Se	82	57.1	5.5	**0.3743**	0.026	6.9	ug/L	15	Standard
	Se-1	77	124.7	7.3	**0.9281**	0.125	13.4	ug/L	48	Standard
>	**Ga**	71	286.7	21.7				mg/L	90	Standard
	Rb	85	31280.6	0.6				ug/L	23	Standard
	Y	89	445698.9	0.7				ug/L	448510	Standard
>	**Rh**	103	16.7	45.8				ug/L	15	Standard
	Mo	98	229.5	3.6	**0.0662**	0.004	6.0	ug/L	11	Standard
	Ag	107	129.3	14.3	**0.0031**	0.003	91.8	ug/L	134	Standard
	Cd	111	79.6	15.3	**0.0346**	0.006	17.8	mg/L	7	Standard
	Cd	114	219.2	32.9	**0.0375**	0.014	38.3	ug/L	16	Standard
>	**In**	115	608521.7	1.3				ug/L	611117	Standard
	Sn	118	713.4	5.3	**0.0192**	0.007	34.1	ug/L	518	Standard
	Sb	123	215.5	21.1	**0.0330**	0.009	27.4	ug/L	127	Standard
	Ba	135	73971.5	1.1	**35.9586**	0.191	0.5	ug/L	45	Standard
	Ce	140	14728.8	3.9				ug/L	298	Standard
>	**Tb**	159	956871.9	1.4				ug/L	977942	Standard
	Ho	165	336.7	6.7				ug/L	12	Standard
	Tl	203	279.3	8.0	**0.0283**	0.003	10.3	ug/L	31	Standard
	Tl	205	295.0	11.9	**0.0379**	0.004	11.8	ug/L	20	Standard
	Pb	206	724.7	2.0	**0.0702**	0.002	2.4	ug/L	286	Standard
	Pb	207	605.7	5.7	**0.0660**	0.008	12.2	ug/L	257	Standard
	Pb	208	2584.1	3.2	**0.0763**	0.006	7.3	ug/L	956	Standard
	U	238	605.0	8.8	**0.0814**	0.008	10.1	ug/L	17	Standard
>	**Bi**	209	558048.4	1.1				ug/L	557080	Standard

Sample ID: L1606079403
Report Date/Time: Tuesday, June 28, 2016 20:38:30
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4734.1	2.7	**6.2297**	0.190	3.1	mg/L	32	Standard
K	39	378.3	13.8	**1.1326**	0.165	14.5	mg/L	30	Standard
Ca	43	55.0	18.2	**3.0097**	1.672	55.5	mg/L	48	Standard
Fe	54	512.5	3.2	**0.3260**	0.027	8.3	mg/L	159	Standard
Fe	57	288.3	19.9	**0.3473**	0.203	58.5	mg/L	195	Standard
> **Sc-1**	45	39648.5	2.7				mg/L	39543	Standard
Cl	35	35720.7	1.2				ug/L	27642	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	10650.4	4.7				ug/L	1647	Standard
P	31	21555.3	1.7				ug/L	15231	Standard
S	34	2313.5	4.4				ug/L	1845	Standard
Sr	88	90.0	14.7				ug/L	70	Standard
C	12	206.7	17.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0					mg/L	0	Standard
Dy	164	611.8	5.5				mg/L	12	Standard
Ho-1	165	336.7	6.7				mg/L	12	Standard
Er	166	310.0	11.6				mg/L	20	Standard
I	127	51010.9	6.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.877	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.489	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606079403
Report Date/Time: Tuesday, June 28, 2016 20:38:30
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	99.575	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	100.174	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606079404
Sample Date/Time: Tuesday, June 28, 2016 20:39:24
Number of Replicates: 3
Autosampler Position: 314
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77373.8	1.6				ug/L	81760	Standard
	Be	9	33.3	45.8	**0.0257**	0.016	61.7	ug/L	17	Standard
	Al	27	6530294.9	1.7	**44.7685**	1.484	3.3	ug/L	1410	Standard
	Sc	45	40462.3	2.9				ug/L	39543	Standard
	Ti	47	1753.8	22.2	**6.9932**	1.564	22.4	ug/L	39	Standard
	V	51	4913.4	4.4	**0.5113**	0.026	5.2	ug/L	1415	Standard
	Cr	52	10044.0	5.5	**0.4080**	0.064	15.7	ug/L	6946	Standard
	Cr	53	2381.9	7.8	**2.4159**	0.236	9.8	ug/L	328	Standard
	Mn	55	8647420.3	0.2	**1354.8198**	14.536	1.1	ug/L	823	Standard
	Co	59	169235.6	0.5	**24.4356**	0.161	0.7	ug/L	252	Standard
	Ni	60	12361.0	0.9	**6.5619**	0.023	0.4	ug/L	192	Standard
	Cu	65	806.0	0.9	**0.3401**	0.009	2.5	ug/L	141	Standard
	Zn	66	7577.2	1.7	**6.1942**	0.082	1.3	ug/L	491	Standard
>	**Ge**	72	507111.0	1.2				ug/L	515087	Standard
	As	75	265.8	16.5	**0.2802**	0.037	13.2	ug/L	20	Standard
	Se	82	64.6	10.0	**0.4504**	0.053	11.9	ug/L	15	Standard
	Se-1	77	107.0	10.8	**0.7228**	0.140	19.4	ug/L	48	Standard
>	**Ga**	71	598.3	9.8				mg/L	90	Standard
	Rb	85	32554.9	1.8				ug/L	23	Standard
	Y	89	447221.5	3.0				ug/L	448510	Standard
>	**Rh**	103	20.0	50.0				ug/L	15	Standard
	Mo	98	99.0	16.4	**0.0174**	0.006	35.0	ug/L	11	Standard
	Ag	107	155.7	70.8	**0.0069**	0.016	237.9	ug/L	134	Standard
	Cd	111	107.8	50.7	**0.0474**	0.026	54.2	mg/L	7	Standard
	Cd	114	265.0	52.4	**0.0460**	0.027	58.4	ug/L	16	Standard
>	**In**	115	613219.8	0.3				ug/L	611117	Standard
	Sn	118	870.0	6.6	**0.0447**	0.010	22.4	ug/L	518	Standard
	Sb	123	214.6	24.5	**0.0325**	0.011	34.6	ug/L	127	Standard
	Ba	135	118463.0	0.8	**57.1584**	0.254	0.4	ug/L	45	Standard
	Ce	140	43559.2	3.6				ug/L	298	Standard
>	**Tb**	159	953430.3	1.1				ug/L	977942	Standard
	Ho	165	631.7	5.3				ug/L	12	Standard
	Tl	203	508.0	24.8	**0.0560**	0.015	27.1	ug/L	31	Standard
	Tl	205	431.7	17.4	**0.0571**	0.010	17.4	ug/L	20	Standard
	Pb	206	1022.0	12.5	**0.1271**	0.024	19.0	ug/L	286	Standard
	Pb	207	848.0	9.6	**0.1173**	0.017	14.2	ug/L	257	Standard
	Pb	208	3317.1	5.1	**0.1168**	0.009	7.7	ug/L	956	Standard
	U	238	210.3	17.6	**0.0266**	0.005	19.3	ug/L	17	Standard
>	**Bi**	209	552632.4	0.7				ug/L	557080	Standard

Sample ID: L1606079404
Report Date/Time: Tuesday, June 28, 2016 20:41:41
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**20.7118**	17.947	86.6	mg/L	0	Standard
Mg	24	5764.4	1.5	**7.4467**	0.270	3.6	mg/L	32	Standard
K	39	563.3	10.2	**1.7170**	0.152	8.8	mg/L	30	Standard
Ca	43	75.0	17.6	**5.8105**	2.161	37.2	mg/L	48	Standard
Fe	54	999.2	5.7	**0.7481**	0.074	9.9	mg/L	159	Standard
Fe	57	490.0	3.5	**0.9936**	0.094	9.5	mg/L	195	Standard
> **Sc-1**	45	40462.3	2.9				mg/L	39543	Standard
Cl	35	35374.6	1.6				ug/L	27642	Standard
Kr	83	2.7	78.1				ug/L	2	Standard
Br	81	12395.1	8.6				ug/L	1647	Standard
P	31	20597.3	1.4				ug/L	15231	Standard
S	34	2416.9	2.7				ug/L	1845	Standard
Sr	88	128.3	19.6				ug/L	70	Standard
C	12	236.7	2.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	1138.1	44.8				mg/L	12	Standard
Ho-1	165	631.7	5.3				mg/L	12	Standard
Er	166	596.7	6.3				mg/L	20	Standard
I	127	203787.2	10.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.635	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.452	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606079404
Report Date/Time: Tuesday, June 28, 2016 20:41:41
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	100.344
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	99.202
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606079404
Report Date/Time: Tuesday, June 28, 2016 20:41:41
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606082202
Sample Date/Time: Tuesday, June 28, 2016 20:42:36
Number of Replicates: 3
Autosampler Position: 315
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79831.5	1.5				ug/L	81760	Standard
	Be	9	18.3	63.0	**0.0091**	0.012	129.5	ug/L	17	Standard
	Al	27	3326140.8	2.2	**22.0876**	0.529	2.4	ug/L	1410	Standard
	Sc	45	40903.5	2.4				ug/L	39543	Standard
	Ti	47	258.3	7.7	**0.8950**	0.111	12.4	ug/L	39	Standard
	V	51	1427.8	26.8	**0.0038**	0.061	1620.7	ug/L	1415	Standard
	Cr	52	8606.8	1.2	**0.2025**	0.023	11.3	ug/L	6946	Standard
	Cr	53	7031.6	3.4	**7.8173**	0.215	2.7	ug/L	328	Standard
	Mn	55	245282.9	2.4	**38.3872**	0.371	1.0	ug/L	823	Standard
	Co	59	360.0	5.7	**0.0208**	0.001	6.7	ug/L	252	Standard
	Ni	60	675.7	7.1	**0.2607**	0.030	11.7	ug/L	192	Standard
	Cu	65	319.3	4.0	**0.0805**	0.011	13.1	ug/L	141	Standard
	Zn	66	1777.8	5.5	**0.8236**	0.060	7.3	ug/L	491	Standard
>	**Ge**	72	505790.0	3.0				ug/L	515087	Standard
	As	75	361.2	27.3	**0.3676**	0.091	24.7	ug/L	20	Standard
	Se	82	151.2	10.8	**1.2184**	0.141	11.6	ug/L	15	Standard
	Se-1	77	286.7	5.2	**3.1548**	0.107	3.4	ug/L	48	Standard
>	**Ga**	71	148.3	15.6				mg/L	90	Standard
	Rb	85	12600.2	4.8				ug/L	23	Standard
	Y	89	440998.1	2.4				ug/L	448510	Standard
>	**Rh**	103	48.3	6.0				ug/L	15	Standard
	Mo	98	329.5	7.2	**0.1031**	0.007	6.5	ug/L	11	Standard
	Ag	107	82.7	3.0	**-0.0039**	0.000	4.5	ug/L	134	Standard
	Cd	111	39.1	14.9	**0.0154**	0.002	16.0	mg/L	7	Standard
	Cd	114	93.5	42.9	**0.0131**	0.008	57.9	ug/L	16	Standard
>	**In**	115	609428.7	1.7				ug/L	611117	Standard
	Sn	118	616.7	6.1	**0.0027**	0.008	296.4	ug/L	518	Standard
	Sb	123	172.8	22.7	**0.0238**	0.008	32.2	ug/L	127	Standard
	Ba	135	142178.1	1.8	**69.0335**	0.137	0.2	ug/L	45	Standard
	Ce	140	378.3	16.5				ug/L	298	Standard
>	**Tb**	159	949587.4	3.3				ug/L	977942	Standard
	Ho	165	25.0	20.0				ug/L	12	Standard
	Tl	203	147.7	1.7	**0.0134**	0.001	4.1	ug/L	31	Standard
	Tl	205	166.7	29.1	**0.0214**	0.007	33.5	ug/L	20	Standard
	Pb	206	309.7	2.3	**-0.0043**	0.001	31.8	ug/L	286	Standard
	Pb	207	262.7	4.1	**-0.0022**	0.002	83.4	ug/L	257	Standard
	Pb	208	1080.7	3.0	**-0.0011**	0.001	98.3	ug/L	956	Standard
	U	238	30.0	28.9	**0.0014**	0.001	94.3	ug/L	17	Standard
>	**Bi**	209	537934.0	1.4				ug/L	557080	Standard

Sample ID: L1606082202
Report Date/Time: Tuesday, June 28, 2016 20:44:53
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.6139**	18.375	173.1	mg/L	0	Standard
Mg	24	32371.2	3.9	**41.6322**	2.388	5.7	mg/L	32	Standard
K	39	260.0	20.1	**0.7087**	0.176	24.8	mg/L	30	Standard
Ca	43	58.3	19.8	**3.1978**	1.533	48.0	mg/L	48	Standard
Fe	54	1294.3	5.1	**0.9965**	0.083	8.3	mg/L	159	Standard
Fe	57	445.0	8.5	**0.8270**	0.105	12.6	mg/L	195	Standard
Sc-1	45	40903.5	2.4				mg/L	39543	Standard
Cl	35	36431.1	1.6				ug/L	27642	Standard
Kr	83	0.3	173.2				ug/L	2	Standard
Br	81	35805.7	4.9				ug/L	1647	Standard
P	31	25079.1	3.1				ug/L	15231	Standard
S	34	2408.5	3.8				ug/L	1845	Standard
Sr	88	206.7	19.6				ug/L	70	Standard
C	12	176.7	45.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	28.1	3.0				mg/L	12	Standard
Ho-1	165	25.0	20.0				mg/L	12	Standard
Er	166	40.0	43.3				mg/L	20	Standard
I	127	201681.5	7.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		97.641	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		98.195	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606082202
Report Date/Time: Tuesday, June 28, 2016 20:44:53
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.724
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.563
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606082203
Sample Date/Time: Tuesday, June 28, 2016 20:45:47
Number of Replicates: 3
Autosampler Position: 316
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90191.0	1.0				ug/L	81760	Standard
	Be	9	23.3	68.9	**0.0116**	0.015	127.2	ug/L	17	Standard
	Al	27	11365406.0	0.9	**66.8332**	1.185	1.8	ug/L	1410	Standard
	Sc	45	43861.8	3.4				ug/L	39543	Standard
	Ti	47	293.7	8.7	**0.9575**	0.101	10.6	ug/L	39	Standard
	V	51	2061.8	11.0	**0.0753**	0.031	41.0	ug/L	1415	Standard
	Cr	52	35560.4	1.1	**3.7982**	0.065	1.7	ug/L	6946	Standard
	Cr	53	13979.8	6.4	**14.7966**	1.011	6.8	ug/L	328	Standard
	Mn	55	581441.7	2.8	**85.0974**	2.668	3.1	ug/L	823	Standard
	Co	59	11104.7	2.7	**1.4708**	0.043	3.0	ug/L	252	Standard
	Ni	60	4756.1	1.3	**2.2951**	0.027	1.2	ug/L	192	Standard
	Cu	65	585.0	3.7	**0.2018**	0.012	5.9	ug/L	141	Standard
	Zn	66	4219.9	2.2	**2.8317**	0.093	3.3	ug/L	491	Standard
>	**Ge**	72	541965.4	0.4				ug/L	515087	Standard
	As	75	1879.3	4.8	**1.6291**	0.081	5.0	ug/L	20	Standard
	Se	82	403.3	6.2	**3.2102**	0.210	6.5	ug/L	15	Standard
	Se-1	77	499.0	5.4	**5.5758**	0.322	5.8	ug/L	48	Standard
>	**Ga**	71	170.0	12.8				mg/L	90	Standard
	Rb	85	898318.0	1.7				ug/L	23	Standard
	Y	89	454033.5	1.6				ug/L	448510	Standard
>	**Rh**	103	158.3	15.6				ug/L	15	Standard
	Mo	98	390.2	3.3	**0.1238**	0.006	4.6	ug/L	11	Standard
	Ag	107	91.7	18.8	**-0.0027**	0.003	94.7	ug/L	134	Standard
	Cd	111	57.3	18.5	**0.0236**	0.005	19.9	mg/L	7	Standard
	Cd	114	162.3	23.5	**0.0261**	0.008	29.2	ug/L	16	Standard
>	**In**	115	617377.1	1.4				ug/L	611117	Standard
	Sn	118	971.7	10.6	**0.0609**	0.019	30.8	ug/L	518	Standard
	Sb	123	275.2	12.0	**0.0450**	0.008	16.8	ug/L	127	Standard
	Ba	135	754414.8	1.4	**361.7131**	4.988	1.4	ug/L	45	Standard
	Ce	140	4038.9	3.4				ug/L	298	Standard
>	**Tb**	159	972701.2	1.2				ug/L	977942	Standard
	Ho	165	60.0					ug/L	12	Standard
	Tl	203	156.0	13.2	**0.0138**	0.002	16.1	ug/L	31	Standard
	Tl	205	116.7	13.1	**0.0137**	0.002	15.9	ug/L	20	Standard
	Pb	206	526.0	2.4	**0.0341**	0.004	10.3	ug/L	286	Standard
	Pb	207	425.7	3.6	**0.0296**	0.004	13.9	ug/L	257	Standard
	Pb	208	1793.4	1.5	**0.0349**	0.002	5.6	ug/L	956	Standard
	U	238	113.7	2.7	**0.0130**	0.001	4.0	ug/L	17	Standard
>	**Bi**	209	555248.6	1.3				ug/L	557080	Standard

Sample ID: L1606082203
Report Date/Time: Tuesday, June 28, 2016 20:48:04
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**9.3851**	16.247	173.1	mg/L	0	Standard
Mg	24	35620.2	1.5	**42.7261**	1.684	3.9	mg/L	32	Standard
K	39	6901.6	2.1	**20.9070**	1.156	5.5	mg/L	30	Standard
Ca	43	215.0	12.3	**24.1234**	3.813	15.8	mg/L	48	Standard
Fe	54	3077.7	3.9	**2.3743**	0.067	2.8	mg/L	159	Standard
Fe	57	1036.7	11.9	**2.5246**	0.275	10.9	mg/L	195	Standard
Sc-1	45	43861.8	3.4				mg/L	39543	Standard
Cl	35	37418.9	4.6				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	90214.6	3.4				ug/L	1647	Standard
P	31	25815.3	2.7				ug/L	15231	Standard
S	34	2615.2	3.8				ug/L	1845	Standard
Sr	88	703.3	5.7				ug/L	70	Standard
C	12	206.7	5.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	96.0	27.7				mg/L	12	Standard
Ho-1	165	60.0					mg/L	12	Standard
Er	166	83.3	6.9				mg/L	20	Standard
I	127	135411.8	5.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.312	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		105.218	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082203
Report Date/Time: Tuesday, June 28, 2016 20:48:04
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	101.024	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	99.671	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Ba 135 Upper, S, EEE	Ba	135	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606082302
Sample Date/Time: Tuesday, June 28, 2016 20:48:59
Number of Replicates: 3
Autosampler Position: 317
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	108277.4	2.9				ug/L	81760	Standard
	Be	9	45.0	44.4	**0.0244**	0.015	61.4	ug/L	17	Standard
	Al	27	155300534.6	2.6	**761.2127**	30.613	4.0	ug/L	1410	Standard
	Sc	45	45277.6	3.5				ug/L	39543	Standard
	Ti	47	403.3	5.8	**1.4273**	0.064	4.5	ug/L	39	Standard
	V	51	3112.6	16.0	**0.2334**	0.070	30.2	ug/L	1415	Standard
	Cr	52	14043.2	7.2	**0.9232**	0.105	11.4	ug/L	6946	Standard
	Cr	53	48930.1	15.5	**54.3111**	7.500	13.8	ug/L	328	Standard
	Mn	55	2748271.5	3.9	**416.1426**	7.847	1.9	ug/L	823	Standard
	Co	59	187853.6	1.8	**26.2312**	0.060	0.2	ug/L	252	Standard
	Ni	60	24191.7	7.1	**12.5042**	0.665	5.3	ug/L	192	Standard
	Cu	65	2608.9	7.7	**1.2568**	0.078	6.2	ug/L	141	Standard
	Zn	66	8434.7	3.5	**6.7297**	0.123	1.8	ug/L	491	Standard
>	**Ge**	72	524401.5	2.0				ug/L	515087	Standard
	As	75	6667.1	7.6	**5.8608**	0.340	5.8	ug/L	20	Standard
	Se	82	2463.0	3.4	**20.8902**	0.307	1.5	ug/L	15	Standard
	Se-1	77	5803.1	9.1	**74.8730**	5.475	7.3	ug/L	48	Standard
>	**Ga**	71	283.3	11.5				mg/L	90	Standard
	Rb	85	14086.5	1.8				ug/L	23	Standard
	Y	89	450377.0	4.3				ug/L	448510	Standard
>	**Rh**	103	1153.4	2.0				ug/L	15	Standard
	Mo	98	228.3	2.1	**0.0678**	0.002	2.3	ug/L	11	Standard
	Ag	107	100.7	7.3	**-0.0008**	0.001	114.3	ug/L	134	Standard
	Cd	111	85.9	11.5	**0.0385**	0.005	14.0	mg/L	7	Standard
	Cd	114	207.5	20.1	**0.0360**	0.007	20.1	ug/L	16	Standard
>	**In**	115	594734.3	2.8				ug/L	611117	Standard
	Sn	118	838.4	7.1	**0.0436**	0.007	15.0	ug/L	518	Standard
	Sb	123	341.7	13.9	**0.0614**	0.008	13.7	ug/L	127	Standard
	Ba	135	7531184.6	2.9	**3748.3398**	4.365	0.1	ug/L	45	Standard
	Ce	140	2656.9	7.0				ug/L	298	Standard
>	**Tb**	159	994456.1	1.2				ug/L	977942	Standard
	Ho	165	153.3	5.0				ug/L	12	Standard
	Tl	203	149.0	11.0	**0.0143**	0.002	16.0	ug/L	31	Standard
	Tl	205	131.7	44.0	**0.0171**	0.008	49.5	ug/L	20	Standard
	Pb	206	558.3	2.7	**0.0476**	0.002	3.6	ug/L	286	Standard
	Pb	207	480.7	4.4	**0.0481**	0.004	8.4	ug/L	257	Standard
	Pb	208	1800.4	1.7	**0.0423**	0.001	2.5	ug/L	956	Standard
	U	238	2848.6	3.5	**0.4250**	0.011	2.7	ug/L	17	Standard
>	**Bi**	209	517504.0	1.7				ug/L	557080	Standard

Sample ID: L1606082302
Report Date/Time: Tuesday, June 28, 2016 20:51:16
Page 1

Approved: June 29, 2016

Na	23	15.0	33.3	**83.8040**	26.058	31.1	mg/L	0	Standard
Mg	24	304856.8	3.2	**354.4066**	5.599	1.6	mg/L	32	Standard
K	39	523.3	14.8	**1.3989**	0.174	12.4	mg/L	30	Standard
Ca	43	1175.0	5.2	**150.6324**	9.407	6.2	mg/L	48	Standard
Fe	54	2706.9	4.6	**2.0025**	0.071	3.5	mg/L	159	Standard
Fe	57	1593.4	6.7	**4.0698**	0.154	3.8	mg/L	195	Standard
> **Sc-1**	45	45277.6	3.5				mg/L	39543	Standard
Cl	35	40644.4	1.3				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	610984.3	3.4				ug/L	1647	Standard
P	31	30180.0	2.7				ug/L	15231	Standard
S	34	2781.9	5.2				ug/L	1845	Standard
Sr	88	4374.0	2.8				ug/L	70	Standard
C	12	443.3	22.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	166.1	26.6				mg/L	12	Standard
Ho-1	165	153.3	5.0				mg/L	12	Standard
Er	166	220.0	28.4				mg/L	20	Standard
I	127	589063.3	13.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.433	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.808	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082302
Report Date/Time: Tuesday, June 28, 2016 20:51:16
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.319
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	92.896
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606082302
Report Date/Time: Tuesday, June 28, 2016 20:51:16
Page 3

Approved: June 29, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606082303
Sample Date/Time: Tuesday, June 28, 2016 20:52:10
Number of Replicates: 3
Autosampler Position: 318
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93244.4	1.1				ug/L	81760	Standard
	Be	9	28.3	10.2	**0.0153**	0.003	16.9	ug/L	17	Standard
	Al	27	94744933.1	0.6	**538.9798**	6.462	1.2	ug/L	1410	Standard
	Sc	45	47113.2	4.5				ug/L	39543	Standard
	Ti	47	553.7	11.7	**2.0175**	0.206	10.2	ug/L	39	Standard
	V	51	4328.2	13.5	**0.4057**	0.092	22.6	ug/L	1415	Standard
	Cr	52	14490.9	2.4	**0.9864**	0.018	1.8	ug/L	6946	Standard
	Cr	53	43709.7	3.6	**48.5418**	1.565	3.2	ug/L	328	Standard
	Mn	55	8319616.3	2.0	**1259.6840**	13.722	1.1	ug/L	823	Standard
	Co	59	19767.9	2.2	**2.7309**	0.056	2.1	ug/L	252	Standard
	Ni	60	10810.2	4.4	**5.5279**	0.127	2.3	ug/L	192	Standard
	Cu	65	2456.2	2.4	**1.1782**	0.021	1.8	ug/L	141	Standard
	Zn	66	5426.0	2.4	**4.0320**	0.062	1.5	ug/L	491	Standard
>	**Ge**	72	524719.7	2.2				ug/L	515087	Standard
	As	75	3752.6	7.8	**3.3171**	0.255	7.7	ug/L	20	Standard
	Se	82	1738.6	3.2	**14.7077**	0.502	3.4	ug/L	15	Standard
	Se-1	77	4761.1	4.1	**61.3297**	2.628	4.3	ug/L	48	Standard
>	**Ga**	71	408.3	8.7				mg/L	90	Standard
	Rb	85	9714.8	3.7				ug/L	23	Standard
	Y	89	460620.0	3.3				ug/L	448510	Standard
>	**Rh**	103	891.7	10.4				ug/L	15	Standard
	Mo	98	389.5	4.4	**0.1260**	0.003	2.4	ug/L	11	Standard
	Ag	107	110.7	7.4	**0.0004**	0.001	272.1	ug/L	134	Standard
	Cd	111	332.6	3.0	**0.1538**	0.007	4.3	mg/L	7	Standard
	Cd	114	853.2	3.8	**0.1605**	0.008	5.1	ug/L	16	Standard
>	**In**	115	607039.3	2.6				ug/L	611117	Standard
	Sn	118	751.7	13.7	**0.0260**	0.016	61.8	ug/L	518	Standard
	Sb	123	376.4	10.5	**0.0674**	0.007	10.4	ug/L	127	Standard
	Ba	135	4357695.0	1.8	**2125.2206**	17.296	0.8	ug/L	45	Standard
	Ce	140	11329.2	2.7				ug/L	298	Standard
>	**Tb**	159	992575.1	2.0				ug/L	977942	Standard
	Ho	165	278.3	9.0				ug/L	12	Standard
	Tl	203	135.0	5.8	**0.0123**	0.001	6.8	ug/L	31	Standard
	Tl	205	110.0	27.6	**0.0138**	0.004	31.4	ug/L	20	Standard
	Pb	206	907.4	3.0	**0.1159**	0.005	4.7	ug/L	286	Standard
	Pb	207	750.0	6.5	**0.1064**	0.009	8.4	ug/L	257	Standard
	Pb	208	3103.1	4.1	**0.1154**	0.005	4.4	ug/L	956	Standard
	U	238	44356.2	2.2	**6.6145**	0.107	1.6	ug/L	17	Standard
>	**Bi**	209	521151.7	1.4				ug/L	557080	Standard

Sample ID: L1606082303
Report Date/Time: Tuesday, June 28, 2016 20:54:27
Page 1

Approved: June 29, 2016

Na	23	10.0	86.6	**54.6526**	47.439	86.8	mg/L	0	Standard
Mg	24	287663.3	3.0	**321.5385**	8.440	2.6	mg/L	32	Standard
K	39	226.7	14.2	**0.5032**	0.104	20.6	mg/L	30	Standard
Ca	43	1070.0	2.9	**131.3674**	10.299	7.8	mg/L	48	Standard
Fe	54	476.6	5.9	**0.2261**	0.038	16.9	mg/L	159	Standard
Fe	57	1008.4	3.2	**2.2367**	0.165	7.4	mg/L	195	Standard
> **Sc-1**	45	47113.2	4.5				mg/L	39543	Standard
Cl	35	40790.8	0.7				ug/L	27642	Standard
Kr	83	0.7	173.2				ug/L	2	Standard
Br	81	436089.5	3.9				ug/L	1647	Standard
P	31	27271.2	4.0				ug/L	15231	Standard
S	34	2836.9	4.2				ug/L	1845	Standard
Sr	88	2816.9	2.8				ug/L	70	Standard
C	12	376.7	12.5				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	20.0	86.6				mg/L	0	Standard
Dy	164	345.0	17.3				mg/L	12	Standard
Ho-1	165	278.3	9.0				mg/L	12	Standard
Er	166	243.3	18.5				mg/L	20	Standard
I	127	900084.0	10.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		114.046	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.870	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082303
Report Date/Time: Tuesday, June 28, 2016 20:54:27
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.333
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		93.551
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606082303
Report Date/Time: Tuesday, June 28, 2016 20:54:27
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606082304
Sample Date/Time: Tuesday, June 28, 2016 20:55:22
Number of Replicates: 3
Autosampler Position: 319
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74476.1	3.3				ug/L	81760	Standard
	Be	9	15.0	115.5	**0.0072**	0.020	276.2	ug/L	17	Standard
	Al	27	20080706.8	2.9	**143.0136**	1.080	0.8	ug/L	1410	Standard
	Sc	45	42894.0	3.7				ug/L	39543	Standard
	Ti	47	325.7	10.6	**1.1562**	0.137	11.8	ug/L	39	Standard
	V	51	1429.9	29.9	**0.0011**	0.063	5474.1	ug/L	1415	Standard
	Cr	52	7720.6	1.3	**0.0621**	0.011	18.4	ug/L	6946	Standard
	Cr	53	9723.1	6.1	**10.8363**	0.676	6.2	ug/L	328	Standard
	Mn	55	4663144.1	0.8	**725.7255**	4.778	0.7	ug/L	823	Standard
	Co	59	1828.4	1.1	**0.2313**	0.003	1.3	ug/L	252	Standard
	Ni	60	1442.4	3.3	**0.6679**	0.026	3.9	ug/L	192	Standard
	Cu	65	719.0	5.1	**0.2910**	0.019	6.5	ug/L	141	Standard
	Zn	66	7525.5	1.3	**6.1013**	0.101	1.7	ug/L	491	Standard
>	**Ge**	72	510419.4	0.4				ug/L	515087	Standard
	As	75	3004.5	4.7	**2.7368**	0.118	4.3	ug/L	20	Standard
	Se	82	527.6	3.3	**4.5049**	0.149	3.3	ug/L	15	Standard
	Se-1	77	877.0	5.5	**11.0280**	0.616	5.6	ug/L	48	Standard
>	**Ga**	71	135.0	44.9				mg/L	90	Standard
	Rb	85	21460.2	1.4				ug/L	23	Standard
	Y	89	449502.6	1.2				ug/L	448510	Standard
>	**Rh**	103	240.0	18.2				ug/L	15	Standard
	Mo	98	71.6	8.8	**0.0080**	0.002	28.5	ug/L	11	Standard
	Ag	107	77.3	9.2	**-0.0044**	0.001	24.3	ug/L	134	Standard
	Cd	111	102.2	7.3	**0.0460**	0.004	8.2	mg/L	7	Standard
	Cd	114	276.1	15.2	**0.0494**	0.008	16.6	ug/L	16	Standard
>	**In**	115	597587.3	0.9				ug/L	611117	Standard
	Sn	118	641.7	5.5	**0.0090**	0.005	57.4	ug/L	518	Standard
	Sb	123	117.8	11.7	**0.0127**	0.003	24.6	ug/L	127	Standard
	Ba	135	1219381.1	1.8	**603.9591**	5.851	1.0	ug/L	45	Standard
	Ce	140	360.0	14.7				ug/L	298	Standard
>	**Tb**	159	948906.4	1.3				ug/L	977942	Standard
	Ho	165	16.7	17.3				ug/L	12	Standard
	Tl	203	82.3	3.1	**0.0057**	0.000	3.2	ug/L	31	Standard
	Tl	205	61.7	24.8	**0.0067**	0.002	33.3	ug/L	20	Standard
	Pb	206	418.0	3.1	**0.0190**	0.003	17.6	ug/L	286	Standard
	Pb	207	378.0	5.3	**0.0247**	0.003	13.2	ug/L	257	Standard
	Pb	208	1532.7	1.1	**0.0263**	0.002	6.1	ug/L	956	Standard
	U	238	34.3	24.8	**0.0022**	0.001	61.6	ug/L	17	Standard
>	**Bi**	209	521633.3	1.5				ug/L	557080	Standard

Sample ID: L1606082304
Report Date/Time: Tuesday, June 28, 2016 20:57:38
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**19.9869**	34.610	173.2	mg/L	0	Standard
Mg	24	52154.6	2.1	**63.9697**	1.039	1.6	mg/L	32	Standard
K	39	548.3	16.2	**1.5625**	0.223	14.3	mg/L	30	Standard
Ca	43	246.7	5.1	**29.2175**	2.054	7.0	mg/L	48	Standard
Fe	54	45285.0	1.2	**37.6782**	1.468	3.9	mg/L	159	Standard
Fe	57	11722.9	3.7	**35.8594**	0.146	0.4	mg/L	195	Standard
> **Sc-1**	45	42894.0	3.7				mg/L	39543	Standard
Cl	35	35015.1	2.5				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	123689.6	1.5				ug/L	1647	Standard
P	31	28713.9	1.8				ug/L	15231	Standard
S	34	2633.6	10.4				ug/L	1845	Standard
Sr	88	703.3	3.4				ug/L	70	Standard
C	12	210.0	12.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	32.1	16.4				mg/L	12	Standard
Ho-1	165	16.7	17.3				mg/L	12	Standard
Er	166	26.7	57.3				mg/L	20	Standard
I	127	194264.8	2.9				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		91.091	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		99.094	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606082304
Report Date/Time: Tuesday, June 28, 2016 20:57:38
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.786
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	93.637
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606082304
Report Date/Time: Tuesday, June 28, 2016 20:57:38
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606084302
Sample Date/Time: Tuesday, June 28, 2016 20:58:33
Number of Replicates: 3
Autosampler Position: 320
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77271.6	2.1				ug/L	81760	Standard
	Be	9	30.0	44.1	**0.0221**	0.014	61.6	ug/L	17	Standard
	Al	27	13023291.9	1.0	**89.3957**	1.109	1.2	ug/L	1410	Standard
	Sc	45	41354.7	3.2				ug/L	39543	Standard
	Ti	47	172.3	20.3	**0.4975**	0.149	29.9	ug/L	39	Standard
	V	51	1869.8	10.2	**0.0496**	0.023	45.4	ug/L	1415	Standard
	Cr	52	9029.4	5.3	**0.1789**	0.049	27.5	ug/L	6946	Standard
	Cr	53	1933.5	16.3	**1.7631**	0.385	21.8	ug/L	328	Standard
	Mn	55	2010517.0	2.2	**295.6150**	12.044	4.1	ug/L	823	Standard
	Co	59	28381.9	0.9	**3.8204**	0.101	2.6	ug/L	252	Standard
	Ni	60	51973.3	1.5	**26.2049**	0.504	1.9	ug/L	192	Standard
	Cu	65	2375.5	2.3	**1.1012**	0.038	3.5	ug/L	141	Standard
	Zn	66	19572.0	0.3	**16.1981**	0.371	2.3	ug/L	491	Standard
>	**Ge**	72	540385.6	2.0				ug/L	515087	Standard
	As	75	349.5	8.0	**0.3364**	0.018	5.2	ug/L	20	Standard
	Se	82	126.5	4.8	**0.9278**	0.047	5.1	ug/L	15	Standard
	Se-1	77	166.7	13.3	**1.3933**	0.317	22.7	ug/L	48	Standard
>	**Ga**	71	106.7	22.2				mg/L	90	Standard
	Rb	85	40220.0	0.6				ug/L	23	Standard
	Y	89	449597.2	1.4				ug/L	448510	Standard
>	**Rh**	103	146.7	12.0				ug/L	15	Standard
	Mo	98	594.8	3.0	**0.2021**	0.008	3.8	ug/L	11	Standard
	Ag	107	92.7	31.3	**-0.0024**	0.004	183.9	ug/L	134	Standard
	Cd	111	161.3	7.4	**0.0729**	0.005	7.5	mg/L	7	Standard
	Cd	114	441.9	5.3	**0.0806**	0.004	5.2	ug/L	16	Standard
>	**In**	115	608035.7	0.5				ug/L	611117	Standard
	Sn	118	676.7	9.6	**0.0131**	0.012	88.5	ug/L	518	Standard
	Sb	123	188.3	22.2	**0.0273**	0.009	32.4	ug/L	127	Standard
	Ba	135	24154.9	0.6	**11.7345**	0.088	0.8	ug/L	45	Standard
	Ce	140	2120.1	7.8				ug/L	298	Standard
>	**Tb**	159	946861.5	2.0				ug/L	977942	Standard
	Ho	165	63.3	29.9				ug/L	12	Standard
	Tl	203	1826.8	0.7	**0.2203**	0.002	1.1	ug/L	31	Standard
	Tl	205	1685.1	3.3	**0.2367**	0.005	2.2	ug/L	20	Standard
	Pb	206	439.0	7.7	**0.0207**	0.006	26.8	ug/L	286	Standard
	Pb	207	381.0	5.7	**0.0231**	0.004	17.6	ug/L	257	Standard
	Pb	208	1509.0	1.3	**0.0226**	0.000	0.8	ug/L	956	Standard
	U	238	1189.4	4.4	**0.1693**	0.007	4.4	ug/L	17	Standard
>	**Bi**	209	536700.0	1.2				ug/L	557080	Standard

Sample ID: L1606084302
Report Date/Time: Tuesday, June 28, 2016 21:00:50
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**30.9676**	31.632	102.1	mg/L	0	Standard
Mg	24	21311.6	1.7	**27.0900**	1.076	4.0	mg/L	32	Standard
K	39	978.4	4.6	**3.0186**	0.057	1.9	mg/L	30	Standard
Ca	43	386.7	7.5	**50.8608**	4.765	9.4	mg/L	48	Standard
Fe	54	204.3	23.4	**0.0400**	0.042	106.1	mg/L	159	Standard
Fe	57	446.7	11.3	**0.8216**	0.208	25.3	mg/L	195	Standard
> **Sc-1**	45	41354.7	3.2				mg/L	39543	Standard
Cl	35	48993.8	3.0				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	7075.0	5.6				ug/L	1647	Standard
P	31	18566.4	1.9				ug/L	15231	Standard
S	34	2263.5	0.1				ug/L	1845	Standard
Sr	88	423.3	3.6				ug/L	70	Standard
C	12	123.3	40.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	104.0	29.5				mg/L	12	Standard
Ho-1	165	63.3	29.9				mg/L	12	Standard
Er	166	56.7	27.0				mg/L	20	Standard
I	127	48509.5	7.5				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		94.510	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		104.912	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606084302
Report Date/Time: Tuesday, June 28, 2016 21:00:50
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.496
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.342
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606084302
Report Date/Time: Tuesday, June 28, 2016 21:00:50
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 21:01:46
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	77410.6	1.1				ug/L	81760	Standard
	Be	9	48694.9	4.0	**51.7876**	2.330	4.5	ug/L	17	Standard
	Al	27	7498260.2	2.7	**51.3685**	1.536	3.0	ug/L	1410	Standard
	Sc	45	39994.4	1.6				ug/L	39543	Standard
	Ti	47	24919.8	1.5	**100.1235**	0.955	1.0	ug/L	39	Standard
	V	51	352037.7	1.5	**50.4818**	0.410	0.8	ug/L	1415	Standard
	Cr	52	359056.3	3.2	**50.5608**	1.258	2.5	ug/L	6946	Standard
	Cr	53	45264.2	3.2	**51.2590**	1.232	2.4	ug/L	328	Standard
	Mn	55	341978.2	1.6	**52.6439**	0.559	1.1	ug/L	823	Standard
	Co	59	358967.4	0.9	**51.1022**	0.460	0.9	ug/L	252	Standard
	Ni	60	95904.1	3.3	**50.8527**	1.284	2.5	ug/L	192	Standard
	Cu	65	96846.7	3.1	**50.9133**	1.132	2.2	ug/L	141	Standard
	Zn	66	56366.4	2.4	**50.6359**	0.804	1.6	ug/L	491	Standard
>	Ge	72	514670.0	0.9				ug/L	515087	Standard
	As	75	56653.0	2.7	**50.4646**	0.946	1.9	ug/L	20	Standard
	Se	82	5738.6	2.7	**49.7611**	0.957	1.9	ug/L	15	Standard
	Se-1	77	3922.2	4.5	**51.3791**	2.097	4.1	ug/L	48	Standard
>	Ga	71	78.3	3.7				mg/L	90	Standard
	Rb	85	913.4	7.9				ug/L	23	Standard
	Y	89	449629.1	4.0				ug/L	448510	Standard
>	Rh	103	26.7	60.3				ug/L	15	Standard
	Mo	98	269192.0	0.5	**99.9193**	1.745	1.7	ug/L	11	Standard
	Ag	107	324356.7	1.6	**48.4664**	0.560	1.2	ug/L	134	Standard
	Cd	111	104800.0	2.8	**49.1868**	1.080	2.2	mg/L	7	Standard
	Cd	114	253765.7	3.7	**48.9997**	1.658	3.4	ug/L	16	Standard
>	In	115	609141.6	2.1				ug/L	611117	Standard
	Sn	118	292076.9	3.3	**49.4485**	0.824	1.7	ug/L	518	Standard
	Sb	123	232366.9	1.6	**49.3998**	0.394	0.8	ug/L	127	Standard
	Ba	135	100964.1	2.8	**49.0412**	1.097	2.2	ug/L	45	Standard
	Ce	140	81.7	25.5				ug/L	298	Standard
>	Tb	159	939122.1	2.3				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	411985.1	1.5	**50.6376**	0.301	0.6	ug/L	31	Standard
	Tl	205	371737.4	2.9	**52.6085**	0.609	1.2	ug/L	20	Standard
	Pb	206	267607.7	0.9	**51.2233**	0.555	1.1	ug/L	286	Standard
	Pb	207	243032.3	0.6	**51.5800**	0.780	1.5	ug/L	257	Standard
	Pb	208	974203.2	1.1	**53.4395**	0.502	0.9	ug/L	956	Standard
	U	238	359093.2	2.3	**51.8989**	0.179	0.3	ug/L	17	Standard
>	Bi	209	537912.3	2.0				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:04:03
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3960.5	5.8	**5.1535**	0.239	4.6	mg/L	32	Standard
K	39	1450.1	5.1	**4.7011**	0.180	3.8	mg/L	30	Standard
Ca	43	56.7	22.2	**3.1676**	1.896	59.9	mg/L	48	Standard
Fe	54	5602.2	1.3	**4.8773**	0.145	3.0	mg/L	159	Standard
Fe	57	1671.8	3.1	**4.9554**	0.149	3.0	mg/L	195	Standard
Sc-1	45	39994.4	1.6				mg/L	39543	Standard
Cl	35	34924.3	5.3				ug/L	27642	Standard
Kr	83	2.3	49.5				ug/L	2	Standard
Br	81	3010.3	3.5				ug/L	1647	Standard
P	31	18342.8	3.8				ug/L	15231	Standard
S	34	2438.5	4.7				ug/L	1845	Standard
Sr	88	55.0	47.2				ug/L	70	Standard
C	12	96.7	23.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	23.2	100.3				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	3.3	173.2				mg/L	20	Standard
I	127	10668.8	10.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	103.575		
Al	27	102.737		
Sc	45			
Ti	47	100.123		
V	51	100.964		
Cr	52	101.122		
Cr	53			
Mn	55	105.288		
Co	59	102.204		
Ni	60	101.705		
Cu	65	101.827		
Zn	66	101.272		
Ge	72		99.919	
As	75	100.929		
Se	82	99.522		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:04:03
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	99.919		
Ag	107	96.933		
Cd	111	98.374		
Cd	114			
In	115			99.677
Sn	118	98.897		
Sb	123	98.800		
Ba	135	98.082		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.275		
Tl	205			
Pb	206			
Pb	207			
Pb	208	106.879		
U	238	103.798		
Bi	209			96.559
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:04:03
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 21:04:57
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	76325.1	1.4				ug/L	81760	Standard
	Be	9	36.7	47.9	**0.0297**	0.018	61.9	ug/L	17	Standard
	Al	27	7433.8	52.1	**0.0370**	0.026	70.5	ug/L	1410	Standard
	Sc	45	39083.7	0.3				ug/L	39543	Standard
	Ti	47	29.0	21.5	**-0.0509**	0.024	47.8	ug/L	39	Standard
	V	51	1130.0	4.7	**-0.0473**	0.008	16.3	ug/L	1415	Standard
	Cr	52	6124.9	1.6	**-0.1950**	0.012	6.3	ug/L	6946	Standard
	Cr	53	786.7	2.0	**0.5367**	0.019	3.5	ug/L	328	Standard
	Mn	55	1151.0	22.6	**0.0255**	0.040	155.2	ug/L	823	Standard
	Co	59	234.7	8.8	**0.0014**	0.003	211.9	ug/L	252	Standard
	Ni	60	158.3	5.7	**-0.0226**	0.005	21.5	ug/L	192	Standard
	Cu	65	164.3	12.7	**-0.0061**	0.011	179.5	ug/L	141	Standard
	Zn	66	470.3	2.4	**-0.4072**	0.010	2.5	ug/L	491	Standard
>	Ge	72	525676.3	0.2				ug/L	515087	Standard
	As	75	-22.3	21.6	**0.0209**	0.004	19.9	ug/L	20	Standard
	Se	82	22.8	31.8	**0.0747**	0.062	82.4	ug/L	15	Standard
	Se-1	77	66.0	18.9	**0.1398**	0.164	117.4	ug/L	48	Standard
>	Ga	71	56.7	10.2				mg/L	90	Standard
	Rb	85	51.7	11.2				ug/L	23	Standard
	Y	89	445574.4	3.2				ug/L	448510	Standard
>	Rh	103	28.3	36.7				ug/L	15	Standard
	Mo	98	286.8	35.9	**0.0884**	0.038	42.5	ug/L	11	Standard
	Ag	107	119.3	19.2	**0.0018**	0.003	188.2	ug/L	134	Standard
	Cd	111	12.1	17.7	**0.0028**	0.001	35.0	mg/L	7	Standard
	Cd	114	42.6	30.2	**0.0034**	0.003	75.4	ug/L	16	Standard
>	In	115	602370.1	0.9				ug/L	611117	Standard
	Sn	118	773.4	17.1	**0.0309**	0.024	76.6	ug/L	518	Standard
	Sb	123	433.3	25.4	**0.0805**	0.024	30.2	ug/L	127	Standard
	Ba	135	290.7	43.2	**0.1174**	0.060	51.4	ug/L	45	Standard
	Ce	140	18.3	56.8				ug/L	298	Standard
>	Tb	159	950012.1	0.7				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	71.7	40.5	**0.0037**	0.004	95.1	ug/L	31	Standard
	Tl	205	85.0	56.1	**0.0094**	0.007	71.5	ug/L	20	Standard
	Pb	206	271.0	4.2	**-0.0136**	0.003	20.1	ug/L	286	Standard
	Pb	207	260.3	8.3	**-0.0047**	0.005	107.3	ug/L	257	Standard
	Pb	208	1029.7	7.7	**-0.0060**	0.005	82.3	ug/L	956	Standard
	U	238	47.7	42.8	**0.0037**	0.003	79.6	ug/L	17	Standard
>	Bi	209	558519.2	1.3				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:07:14
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	48.3	11.9	**0.0079**	0.008	95.2	mg/L	32	Standard
K	39	18.3	56.8	**-0.0788**	0.036	45.2	mg/L	30	Standard
Ca	43	31.7	24.1	**-0.4803**	1.172	244.0	mg/L	48	Standard
Fe	54	125.9	16.2	**-0.0218**	0.019	87.4	mg/L	159	Standard
Fe	57	238.3	12.6	**0.1894**	0.102	54.0	mg/L	195	Standard
Sc-1	45	39083.7	0.3				mg/L	39543	Standard
Cl	35	33723.5	2.9				ug/L	27642	Standard
Kr	83	2.0	100.0				ug/L	2	Standard
Br	81	2386.9	5.5				ug/L	1647	Standard
P	31	18112.5	4.7				ug/L	15231	Standard
S	34	2391.9	4.0				ug/L	1845	Standard
Sr	88	68.3	29.6				ug/L	70	Standard
C	12	100.0	36.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	18.9	54.4				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	23.3	65.5				mg/L	20	Standard
I	127	7351.8	8.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		102.056	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:07:14
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.569
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	100.258
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:07:14
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141001
Sample Date/Time: Tuesday, June 28, 2016 21:08:10
Number of Replicates: 3
Autosampler Position: 321
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79030.6	2.3				ug/L	81760	Standard
	Be	9	11.7	89.2	**0.0026**	0.011	425.8	ug/L	17	Standard
	Al	27	44776805.2	0.7	**300.5948**	6.187	2.1	ug/L	1410	Standard
	Sc	45	43343.6	2.1				ug/L	39543	Standard
	Ti	47	192.3	8.0	**0.6107**	0.056	9.2	ug/L	39	Standard
	V	51	4772.2	4.6	**0.4823**	0.029	6.1	ug/L	1415	Standard
	Cr	52	9840.2	0.5	**0.3614**	0.015	4.2	ug/L	6946	Standard
	Cr	53	1230.1	5.5	**1.0645**	0.072	6.7	ug/L	328	Standard
	Mn	55	4769620.0	0.5	**738.1503**	9.110	1.2	ug/L	823	Standard
	Co	59	9207.1	1.6	**1.2837**	0.028	2.2	ug/L	252	Standard
	Ni	60	5238.9	2.1	**2.6865**	0.062	2.3	ug/L	192	Standard
	Cu	65	718.7	4.1	**0.2886**	0.013	4.4	ug/L	141	Standard
	Zn	66	2924.3	0.9	**1.8496**	0.029	1.6	ug/L	491	Standard
>	**Ge**	72	513321.3	0.8				ug/L	515087	Standard
	As	75	4203.4	1.1	**3.7919**	0.050	1.3	ug/L	20	Standard
	Se	82	127.5	8.3	**0.9927**	0.098	9.9	ug/L	15	Standard
	Se-1	77	88.3	31.6	**0.4581**	0.375	81.9	ug/L	48	Standard
>	**Ga**	71	145.0	24.1				mg/L	90	Standard
	Rb	85	2930.3	3.3				ug/L	23	Standard
	Y	89	443412.2	1.4				ug/L	448510	Standard
>	**Rh**	103	140.0	15.6				ug/L	15	Standard
	Mo	98	1185.9	4.8	**0.4170**	0.022	5.4	ug/L	11	Standard
	Ag	107	98.0	15.7	**-0.0017**	0.002	128.2	ug/L	134	Standard
	Cd	111	19.9	13.4	**0.0063**	0.001	18.0	mg/L	7	Standard
	Cd	114	87.1	38.1	**0.0117**	0.006	53.6	ug/L	16	Standard
>	**In**	115	614969.4	1.2				ug/L	611117	Standard
	Sn	118	790.0	17.2	**0.0309**	0.023	75.8	ug/L	518	Standard
	Sb	123	280.0	20.4	**0.0462**	0.012	26.5	ug/L	127	Standard
	Ba	135	192282.6	1.4	**92.5375**	1.689	1.8	ug/L	45	Standard
	Ce	140	2006.8	4.1				ug/L	298	Standard
>	**Tb**	159	961897.3	2.5				ug/L	977942	Standard
	Ho	165	81.7	52.8				ug/L	12	Standard
	Tl	203	119.7	8.5	**0.0099**	0.002	15.4	ug/L	31	Standard
	Tl	205	98.3	15.5	**0.0116**	0.002	19.8	ug/L	20	Standard
	Pb	206	482.7	4.9	**0.0287**	0.002	8.4	ug/L	286	Standard
	Pb	207	424.0	4.8	**0.0319**	0.005	14.5	ug/L	257	Standard
	Pb	208	1694.0	1.8	**0.0325**	0.004	11.5	ug/L	956	Standard
	U	238	15563.7	1.1	**2.2446**	0.071	3.2	ug/L	17	Standard
>	**Bi**	209	538649.3	2.3				ug/L	557080	Standard

Sample ID: L1606141001
Report Date/Time: Tuesday, June 28, 2016 21:10:27
Page 1

Approved: June 29, 2016

Na	23	16.7	75.5	**98.5166**	75.230	76.4	mg/L	0	Standard
Mg	24	9014.3	3.2	**10.8981**	0.573	5.3	mg/L	32	Standard
K	39	116.7	19.3	**0.2184**	0.072	32.8	mg/L	30	Standard
Ca	43	525.0	6.7	**67.4527**	5.707	8.5	mg/L	48	Standard
Fe	54	1820.8	2.9	**1.3668**	0.058	4.2	mg/L	159	Standard
Fe	57	973.4	7.5	**2.3749**	0.248	10.4	mg/L	195	Standard
Sc-1	45	43343.6	2.1				mg/L	39543	Standard
Cl	35	38331.8	2.6				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	27825.5	3.4				ug/L	1647	Standard
P	31	23363.0	0.7				ug/L	15231	Standard
S	34	2495.2	2.5				ug/L	1845	Standard
Sr	88	533.3	5.7				ug/L	70	Standard
C	12	353.3	5.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	83.6	23.1				mg/L	12	Standard
Ho-1	165	81.7	52.8				mg/L	12	Standard
Er	166	63.3	50.8				mg/L	20	Standard
I	127	532195.8	11.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		96.662	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.657	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606141001
Report Date/Time: Tuesday, June 28, 2016 21:10:27
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	100.630	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	96.691	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141201
Sample Date/Time: Tuesday, June 28, 2016 21:11:22
Number of Replicates: 3
Autosampler Position: 322
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79318.7	0.7				ug/L	81760	Standard
	Be	9	16.7	45.8	**0.0076**	0.008	106.2	ug/L	17	Standard
	Al	27	14913651.7	1.4	**99.7256**	2.143	2.1	ug/L	1410	Standard
	Sc	45	40395.5	2.6				ug/L	39543	Standard
	Ti	47	500.3	4.7	**1.8979**	0.055	2.9	ug/L	39	Standard
	V	51	36055.9	2.6	**5.1139**	0.040	0.8	ug/L	1415	Standard
	Cr	52	13456.3	2.4	**0.9269**	0.086	9.3	ug/L	6946	Standard
	Cr	53	9844.9	11.5	**11.1480**	1.101	9.9	ug/L	328	Standard
	Mn	55	27340.3	1.0	**4.1779**	0.048	1.1	ug/L	823	Standard
	Co	59	924.7	1.8	**0.1037**	0.005	4.9	ug/L	252	Standard
	Ni	60	2165.5	2.9	**1.0744**	0.025	2.4	ug/L	192	Standard
	Cu	65	1308.7	4.1	**0.6156**	0.023	3.7	ug/L	141	Standard
	Zn	66	7243.7	2.3	**5.9517**	0.120	2.0	ug/L	491	Standard
>	**Ge**	72	502171.0	2.0				ug/L	515087	Standard
	As	75	1256.7	3.1	**1.1867**	0.017	1.5	ug/L	20	Standard
	Se	82	736.7	2.2	**6.4443**	0.037	0.6	ug/L	15	Standard
	Se-1	77	926.0	3.3	**11.8887**	0.285	2.4	ug/L	48	Standard
>	**Ga**	71	176.7	6.5				mg/L	90	Standard
	Rb	85	2561.9	8.5				ug/L	23	Standard
	Y	89	439495.0	0.9				ug/L	448510	Standard
>	**Rh**	103	90.0	38.9				ug/L	15	Standard
	Mo	98	758.7	2.1	**0.2744**	0.005	1.7	ug/L	11	Standard
	Ag	107	112.7	8.8	**0.0013**	0.001	110.1	ug/L	134	Standard
	Cd	111	18.4	13.8	**0.0061**	0.001	20.0	mg/L	7	Standard
	Cd	114	55.1	29.0	**0.0062**	0.003	52.0	ug/L	16	Standard
>	**In**	115	584464.7	0.6				ug/L	611117	Standard
	Sn	118	741.7	7.5	**0.0292**	0.010	34.2	ug/L	518	Standard
	Sb	123	322.5	19.3	**0.0586**	0.014	23.6	ug/L	127	Standard
	Ba	135	23648.1	1.5	**11.9524**	0.216	1.8	ug/L	45	Standard
	Ce	140	4514.0	4.6				ug/L	298	Standard
>	**Tb**	159	922323.8	0.6				ug/L	977942	Standard
	Ho	165	91.7	12.6				ug/L	12	Standard
	Tl	203	130.7	10.2	**0.0124**	0.002	15.2	ug/L	31	Standard
	Tl	205	113.3	46.8	**0.0149**	0.008	53.9	ug/L	20	Standard
	Pb	206	585.3	3.6	**0.0562**	0.004	7.1	ug/L	286	Standard
	Pb	207	533.0	4.2	**0.0629**	0.005	7.8	ug/L	257	Standard
	Pb	208	2107.4	5.4	**0.0632**	0.008	12.1	ug/L	956	Standard
	U	238	32845.9	1.3	**5.0682**	0.050	1.0	ug/L	17	Standard
>	**Bi**	209	503585.9	0.8				ug/L	557080	Standard

Sample ID: L1606141201
Report Date/Time: Tuesday, June 28, 2016 21:13:39
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	118021.2	0.8	**153.7808**	2.811	1.8	mg/L	32	Standard
K	39	101.7	17.3	**0.1948**	0.058	29.7	mg/L	30	Standard
Ca	43	335.0	15.7	**44.3505**	6.519	14.7	mg/L	48	Standard
Fe	54	237.2	9.3	**0.0729**	0.016	22.5	mg/L	159	Standard
Fe	57	498.3	12.0	**1.0226**	0.198	19.4	mg/L	195	Standard
> **Sc-1**	45	40395.5	2.6				mg/L	39543	Standard
Cl	35	43781.2	1.0				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	58978.2	6.0				ug/L	1647	Standard
P	31	22017.7	1.7				ug/L	15231	Standard
S	34	2351.8	1.8				ug/L	1845	Standard
Sr	88	240.0	11.0				ug/L	70	Standard
C	12	356.7	26.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	169.3	7.0				mg/L	12	Standard
Ho-1	165	91.7	12.6				mg/L	12	Standard
Er	166	83.3	13.9				mg/L	20	Standard
I	127	31387.7	30.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		97.014	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.492	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606141201
Report Date/Time: Tuesday, June 28, 2016 21:13:39
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.639
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		90.397
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141202
Sample Date/Time: Tuesday, June 28, 2016 21:14:33
Number of Replicates: 3
Autosampler Position: 323
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83650.3	1.0				ug/L	81760	Standard
	Be	9	21.7	53.3	**0.0115**	0.011	97.0	ug/L	17	Standard
	Al	27	38632215.5	2.3	**244.9578**	5.868	2.4	ug/L	1410	Standard
	Sc	45	43420.5	1.6				ug/L	39543	Standard
	Ti	47	186.0	10.8	**0.5354**	0.077	14.4	ug/L	39	Standard
	V	51	26880.1	4.2	**3.4128**	0.061	1.8	ug/L	1415	Standard
	Cr	52	10918.6	2.1	**0.4115**	0.017	4.1	ug/L	6946	Standard
	Cr	53	7481.8	4.1	**7.6462**	0.526	6.9	ug/L	328	Standard
	Mn	55	33664.4	1.6	**4.7149**	0.144	3.1	ug/L	823	Standard
	Co	59	762.7	5.1	**0.0704**	0.007	10.1	ug/L	252	Standard
	Ni	60	2375.2	1.6	**1.0759**	0.011	1.0	ug/L	192	Standard
	Cu	65	711.7	7.3	**0.2600**	0.027	10.4	ug/L	141	Standard
	Zn	66	3134.0	1.7	**1.8510**	0.112	6.1	ug/L	491	Standard
>	**Ge**	72	550214.2	2.6				ug/L	515087	Standard
	As	75	794.1	8.2	**0.7027**	0.072	10.2	ug/L	20	Standard
	Se	82	234.2	3.9	**1.7863**	0.091	5.1	ug/L	15	Standard
	Se-1	77	482.0	4.4	**5.2700**	0.176	3.3	ug/L	48	Standard
>	**Ga**	71	80.0	16.5				mg/L	90	Standard
	Rb	85	1268.4	2.0				ug/L	23	Standard
	Y	89	467702.3	2.0				ug/L	448510	Standard
>	**Rh**	103	125.0	4.0				ug/L	15	Standard
	Mo	98	395.0	8.4	**0.1254**	0.010	8.1	ug/L	11	Standard
	Ag	107	91.7	6.0	**-0.0027**	0.001	31.4	ug/L	134	Standard
	Cd	111	17.7	18.1	**0.0053**	0.002	30.5	mg/L	7	Standard
	Cd	114	70.4	32.3	**0.0084**	0.004	47.5	ug/L	16	Standard
>	**In**	115	618005.7	2.3				ug/L	611117	Standard
	Sn	118	668.3	14.3	**0.0101**	0.018	180.4	ug/L	518	Standard
	Sb	123	172.5	13.4	**0.0233**	0.005	22.3	ug/L	127	Standard
	Ba	135	80657.5	2.0	**38.6136**	0.662	1.7	ug/L	45	Standard
	Ce	140	231.7	23.7				ug/L	298	Standard
>	**Tb**	159	976289.5	2.7				ug/L	977942	Standard
	Ho	165	20.0	50.0				ug/L	12	Standard
	Tl	203	130.7	0.4	**0.0112**	0.000	4.3	ug/L	31	Standard
	Tl	205	103.3	15.6	**0.0123**	0.003	20.6	ug/L	20	Standard
	Pb	206	343.3	3.7	**0.0019**	0.004	206.6	ug/L	286	Standard
	Pb	207	309.0	3.8	**0.0073**	0.004	55.0	ug/L	257	Standard
	Pb	208	1267.4	6.3	**0.0088**	0.006	66.3	ug/L	956	Standard
	U	238	23437.4	1.1	**3.3661**	0.068	2.0	ug/L	17	Standard
>	**Bi**	209	541066.6	2.7				ug/L	557080	Standard

Sample ID: L1606141202
Report Date/Time: Tuesday, June 28, 2016 21:16:50
Page 1

Approved: June 29, 2016

Na	23	10.0	100.0	**57.8886**	57.492	99.3	mg/L	0	Standard
Mg	24	42571.4	0.8	**51.5637**	1.207	2.3	mg/L	32	Standard
K	39	90.0	30.9	**0.1357**	0.087	64.2	mg/L	30	Standard
Ca	43	438.3	26.2	**55.1842**	15.190	27.5	mg/L	48	Standard
Fe	54	177.1	8.1	**0.0090**	0.012	139.0	mg/L	159	Standard
Fe	57	461.7	5.0	**0.7949**	0.071	8.9	mg/L	195	Standard
Sc-1	45	43420.5	1.6				mg/L	39543	Standard
Cl	35	42170.0	2.1				ug/L	27642	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	34382.4	5.6				ug/L	1647	Standard
P	31	23767.0	4.2				ug/L	15231	Standard
S	34	2608.6	3.1				ug/L	1845	Standard
Sr	88	431.7	5.4				ug/L	70	Standard
C	12	273.3	10.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	32.2	46.2				mg/L	12	Standard
Ho-1	165	20.0	50.0				mg/L	12	Standard
Er	166	23.3	89.2				mg/L	20	Standard
I	127	35633.6	4.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		102.312	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		106.820	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606141202
Report Date/Time: Tuesday, June 28, 2016 21:16:50
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	101.127	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	97.125	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606141202
Report Date/Time: Tuesday, June 28, 2016 21:16:50
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141301
Sample Date/Time: Tuesday, June 28, 2016 21:17:45
Number of Replicates: 3
Autosampler Position: 324
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79916.9	1.5				ug/L	81760	Standard
	Be	9	28.3	27.0	**0.0194**	0.007	38.6	ug/L	17	Standard
	Al	27	38023288.5	2.1	**252.3385**	1.857	0.7	ug/L	1410	Standard
	Sc	45	42571.5	4.4				ug/L	39543	Standard
	Ti	47	396.7	7.8	**1.4239**	0.097	6.8	ug/L	39	Standard
	V	51	6119.0	6.6	**0.6719**	0.073	10.9	ug/L	1415	Standard
	Cr	52	11441.3	0.3	**0.5818**	0.031	5.4	ug/L	6946	Standard
	Cr	53	9654.8	11.4	**10.6403**	1.449	13.6	ug/L	328	Standard
	Mn	55	1634887.0	2.2	**251.2478**	2.194	0.9	ug/L	823	Standard
	Co	59	3215.3	1.8	**0.4249**	0.007	1.5	ug/L	252	Standard
	Ni	60	4170.9	0.8	**2.1033**	0.039	1.9	ug/L	192	Standard
	Cu	65	1200.4	3.8	**0.5391**	0.026	4.9	ug/L	141	Standard
	Zn	66	3234.7	1.4	**2.1147**	0.035	1.7	ug/L	491	Standard
>	**Ge**	72	516685.2	1.8				ug/L	515087	Standard
	As	75	16545.6	0.8	**14.7132**	0.205	1.4	ug/L	20	Standard
	Se	82	192.8	5.0	**1.5497**	0.062	4.0	ug/L	15	Standard
	Se-1	77	648.0	6.7	**7.8638**	0.716	9.1	ug/L	48	Standard
>	**Ga**	71	125.0	18.3				mg/L	90	Standard
	Rb	85	1723.4	7.4				ug/L	23	Standard
	Y	89	425819.3	3.1				ug/L	448510	Standard
>	**Rh**	103	118.3	17.1				ug/L	15	Standard
	Mo	98	4050.2	1.8	**1.5330**	0.030	2.0	ug/L	11	Standard
	Ag	107	94.7	12.2	**-0.0016**	0.002	127.1	ug/L	134	Standard
	Cd	111	117.7	9.5	**0.0542**	0.007	12.3	mg/L	7	Standard
	Cd	114	297.9	4.3	**0.0546**	0.005	8.5	ug/L	16	Standard
>	**In**	115	590223.7	3.4				ug/L	611117	Standard
	Sn	118	1010.0	10.9	**0.0746**	0.013	17.7	ug/L	518	Standard
	Sb	123	1963.6	1.5	**0.4183**	0.013	3.0	ug/L	127	Standard
	Ba	135	194231.1	1.1	**97.4394**	2.308	2.4	ug/L	45	Standard
	Ce	140	3857.2	6.5				ug/L	298	Standard
>	**Tb**	159	963155.2	4.7				ug/L	977942	Standard
	Ho	165	81.7	25.5				ug/L	12	Standard
	Tl	203	50.3	12.9	**0.0015**	0.001	66.6	ug/L	31	Standard
	Tl	205	46.7	32.7	**0.0044**	0.002	52.1	ug/L	20	Standard
	Pb	206	2455.5	1.0	**0.4110**	0.018	4.4	ug/L	286	Standard
	Pb	207	2061.1	1.1	**0.3836**	0.016	4.1	ug/L	257	Standard
	Pb	208	8494.6	1.5	**0.4098**	0.011	2.7	ug/L	956	Standard
	U	238	9671.1	1.3	**1.4065**	0.038	2.7	ug/L	17	Standard
>	**Bi**	209	533828.4	3.8				ug/L	557080	Standard

Sample ID: L1606141301
Report Date/Time: Tuesday, June 28, 2016 21:20:02
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	104970.9	3.0	**129.8006**	2.883	2.2	mg/L	32	Standard
K	39	241.7	3.2	**0.6187**	0.058	9.4	mg/L	30	Standard
Ca	43	386.7	15.6	**49.0389**	6.282	12.8	mg/L	48	Standard
Fe	54	7303.8	3.7	**6.0094**	0.321	5.3	mg/L	159	Standard
Fe	57	2221.8	6.0	**6.3400**	0.172	2.7	mg/L	195	Standard
> **Sc-1**	45	42571.5	4.4				mg/L	39543	Standard
Cl	35	37833.2	3.5				ug/L	27642	Standard
Kr	83	4.7	44.6				ug/L	2	Standard
Br	81	42533.0	4.1				ug/L	1647	Standard
P	31	23852.1	1.0				ug/L	15231	Standard
S	34	2610.2	4.2				ug/L	1845	Standard
Sr	88	403.3	5.0				ug/L	70	Standard
C	12	1170.0	10.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	86.6				mg/L	0	Standard
Dy	164	115.7	9.7				mg/L	12	Standard
Ho-1	165	81.7	25.5				mg/L	12	Standard
Er	166	90.0	29.4				mg/L	20	Standard
I	127	137002.4	10.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		97.746	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.310	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606141301
Report Date/Time: Tuesday, June 28, 2016 21:20:02
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	96.581
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.826
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606141301
Report Date/Time: Tuesday, June 28, 2016 21:20:02
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141601
Sample Date/Time: Tuesday, June 28, 2016 21:20:56
Number of Replicates: 3
Autosampler Position: 325
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92372.4	1.4				ug/L	81760	Standard
	Be	9	16.7	45.8	**0.0051**	0.007	132.4	ug/L	17	Standard
	Al	27	74980741.0	2.1	**430.5985**	11.194	2.6	ug/L	1410	Standard
	Sc	45	42273.9	1.4				ug/L	39543	Standard
	Ti	47	206.0	5.5	**0.6860**	0.049	7.2	ug/L	39	Standard
	V	51	76665.0	2.1	**11.1242**	0.138	1.2	ug/L	1415	Standard
	Cr	52	20589.0	2.5	**1.9813**	0.052	2.6	ug/L	6946	Standard
	Cr	53	16427.3	7.3	**18.8858**	1.445	7.7	ug/L	328	Standard
	Mn	55	4403.6	9.9	**0.5498**	0.075	13.7	ug/L	823	Standard
	Co	59	1375.4	0.7	**0.1698**	0.002	1.0	ug/L	252	Standard
	Ni	60	4849.1	3.4	**2.5400**	0.080	3.2	ug/L	192	Standard
	Cu	65	2467.2	2.0	**1.2432**	0.032	2.6	ug/L	141	Standard
	Zn	66	3268.7	3.6	**2.2359**	0.099	4.4	ug/L	491	Standard
>	**Ge**	72	501391.0	1.2				ug/L	515087	Standard
	As	75	1723.8	1.9	**1.6153**	0.015	1.0	ug/L	20	Standard
	Se	82	1818.6	2.0	**16.1069**	0.131	0.8	ug/L	15	Standard
	Se-1	77	2238.8	2.0	**29.8099**	0.421	1.4	ug/L	48	Standard
>	**Ga**	71	70.0	35.7				mg/L	90	Standard
	Rb	85	6024.5	1.3				ug/L	23	Standard
	Y	89	451516.3	2.4				ug/L	448510	Standard
>	**Rh**	103	580.0	3.0				ug/L	15	Standard
	Mo	98	4894.1	3.5	**1.7784**	0.027	1.5	ug/L	11	Standard
	Ag	107	107.3	12.0	**-0.0004**	0.002	428.3	ug/L	134	Standard
	Cd	111	61.6	14.8	**0.0258**	0.005	19.6	mg/L	7	Standard
	Cd	114	169.0	6.0	**0.0274**	0.002	5.8	ug/L	16	Standard
>	**In**	115	615602.8	3.1				ug/L	611117	Standard
	Sn	118	783.4	1.3	**0.0296**	0.003	11.3	ug/L	518	Standard
	Sb	123	203.3	3.5	**0.0299**	0.001	4.8	ug/L	127	Standard
	Ba	135	8199.6	0.5	**3.9202**	0.126	3.2	ug/L	45	Standard
	Ce	140	360.0	9.7				ug/L	298	Standard
>	**Tb**	159	979932.9	1.3				ug/L	977942	Standard
	Ho	165	28.3	20.4				ug/L	12	Standard
	Tl	203	89.3	17.1	**0.0065**	0.002	28.6	ug/L	31	Standard
	Tl	205	80.0	25.0	**0.0093**	0.003	29.8	ug/L	20	Standard
	Pb	206	345.7	2.0	**0.0043**	0.002	47.1	ug/L	286	Standard
	Pb	207	309.7	5.2	**0.0094**	0.004	39.7	ug/L	257	Standard
	Pb	208	1290.0	0.3	**0.0121**	0.001	10.3	ug/L	956	Standard
	U	238	27903.3	2.3	**4.1288**	0.057	1.4	ug/L	17	Standard
>	**Bi**	209	525055.5	1.5				ug/L	557080	Standard

Sample ID: L1606141601
Report Date/Time: Tuesday, June 28, 2016 21:23:13
Page 1

Approved: June 29, 2016

Na	23	21.7	48.0	**130.4724**	63.085	48.4	mg/L	0	Standard
Mg	24	101248.9	2.2	**126.0185**	2.508	2.0	mg/L	32	Standard
K	39	313.3	18.1	**0.8505**	0.192	22.6	mg/L	30	Standard
Ca	43	933.4	7.3	**127.1886**	7.858	6.2	mg/L	48	Standard
Fe	54	133.4	13.8	**-0.0242**	0.015	63.9	mg/L	159	Standard
Fe	57	858.4	5.3	**2.0873**	0.173	8.3	mg/L	195	Standard
> **Sc-1**	45	42273.9	1.4				mg/L	39543	Standard
Cl	35	65275.0	2.4				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	32608.5	7.2				ug/L	1647	Standard
P	31	23807.1	4.7				ug/L	15231	Standard
S	34	2403.5	5.0				ug/L	1845	Standard
Sr	88	2410.2	4.8				ug/L	70	Standard
C	12	193.3	34.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	0	Standard
Dy	164	21.1	69.3				mg/L	12	Standard
Ho-1	165	28.3	20.4				mg/L	12	Standard
Er	166	46.7	53.9				mg/L	20	Standard
I	127	42467.9	5.2				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		112.980	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		97.341	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606141601
Report Date/Time: Tuesday, June 28, 2016 21:23:13
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	100.734	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	94.251	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606141601
Report Date/Time: Tuesday, June 28, 2016 21:23:13
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141602
Sample Date/Time: Tuesday, June 28, 2016 21:24:08
Number of Replicates: 3
Autosampler Position: 326
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88434.1	2.2				ug/L	81760	Standard
	Be	9	13.3	57.3	**0.0028**	0.007	262.6	ug/L	17	Standard
	Al	27	52144460.9	0.4	**312.8622**	8.112	2.6	ug/L	1410	Standard
	Sc	45	37740.3	1.5				ug/L	39543	Standard
	Ti	47	227.7	10.0	**0.8076**	0.101	12.5	ug/L	39	Standard
	V	51	38489.9	1.8	**5.6761**	0.140	2.5	ug/L	1415	Standard
	Cr	52	19640.1	1.0	**1.9398**	0.010	0.5	ug/L	6946	Standard
	Cr	53	17715.4	1.9	**21.0942**	0.301	1.4	ug/L	328	Standard
	Mn	55	4373.0	6.9	**0.5679**	0.048	8.4	ug/L	823	Standard
	Co	59	870.4	1.6	**0.1002**	0.001	1.2	ug/L	252	Standard
	Ni	60	2666.6	2.1	**1.3989**	0.023	1.6	ug/L	192	Standard
	Cu	65	1074.7	4.8	**0.5100**	0.030	5.9	ug/L	141	Standard
	Zn	66	2890.6	2.0	**1.9736**	0.048	2.4	ug/L	491	Standard
>	**Ge**	72	484924.7	0.7				ug/L	515087	Standard
	As	75	2458.8	5.8	**2.3632**	0.133	5.6	ug/L	20	Standard
	Se	82	892.7	6.7	**8.1181**	0.571	7.0	ug/L	15	Standard
	Se-1	77	1757.1	0.5	**24.0557**	0.109	0.5	ug/L	48	Standard
>	**Ga**	71	76.7	13.6				mg/L	90	Standard
	Rb	85	4900.8	2.1				ug/L	23	Standard
	Y	89	417881.5	5.4				ug/L	448510	Standard
>	**Rh**	103	255.0	18.7				ug/L	15	Standard
	Mo	98	772.2	2.1	**0.2950**	0.006	2.2	ug/L	11	Standard
	Ag	107	68.7	21.6	**-0.0049**	0.003	53.5	ug/L	134	Standard
	Cd	111	15.0	33.1	**0.0048**	0.003	56.9	mg/L	7	Standard
	Cd	114	24.6	131.5	**0.0004**	0.007	1725.2	ug/L	16	Standard
>	**In**	115	555990.7	3.0				ug/L	611117	Standard
	Sn	118	696.7	10.1	**0.0278**	0.015	54.6	ug/L	518	Standard
	Sb	123	106.5	16.8	**0.0120**	0.005	38.5	ug/L	127	Standard
	Ba	135	12524.8	3.0	**6.6489**	0.344	5.2	ug/L	45	Standard
	Ce	140	808.4	5.3				ug/L	298	Standard
>	**Tb**	159	897664.5	1.7				ug/L	977942	Standard
	Ho	165	23.3	68.9				ug/L	12	Standard
	Tl	203	89.0	17.6	**0.0071**	0.002	31.3	ug/L	31	Standard
	Tl	205	90.0	0.0	**0.0115**	0.000	1.5	ug/L	20	Standard
	Pb	206	365.7	4.1	**0.0124**	0.004	32.4	ug/L	286	Standard
	Pb	207	325.3	1.4	**0.0169**	0.000	0.6	ug/L	257	Standard
	Pb	208	1289.4	1.1	**0.0163**	0.000	2.7	ug/L	956	Standard
	U	238	25222.0	1.6	**3.9519**	0.013	0.3	ug/L	17	Standard
>	**Bi**	209	495840.1	1.2				ug/L	557080	Standard

Sample ID: L1606141602
Report Date/Time: Tuesday, June 28, 2016 21:26:25
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**22.3577**	19.364	86.6	mg/L	0	Standard
Mg	24	65916.4	2.1	**91.8901**	2.205	2.4	mg/L	32	Standard
K	39	221.7	4.7	**0.6432**	0.031	4.9	mg/L	30	Standard
Ca	43	530.0	9.4	**79.0853**	9.116	11.5	mg/L	48	Standard
Fe	54	138.0	10.0	**-0.0062**	0.012	195.9	mg/L	159	Standard
Fe	57	510.0	5.1	**1.1793**	0.086	7.3	mg/L	195	Standard
> **Sc-1**	45	37740.3	1.5				mg/L	39543	Standard
Cl	35	44130.9	1.6				ug/L	27642	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	28483.4	4.0				ug/L	1647	Standard
P	31	20263.6	4.9				ug/L	15231	Standard
S	34	2148.5	1.6				ug/L	1845	Standard
Sr	88	1128.4	6.6				ug/L	70	Standard
C	12	163.3	25.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	36.0	33.6				mg/L	12	Standard
Ho-1	165	23.3	68.9				mg/L	12	Standard
Er	166	13.3	114.6				mg/L	20	Standard
I	127	19656.1	3.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		108.163	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.144	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606141602
Report Date/Time: Tuesday, June 28, 2016 21:26:25
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		90.979
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		89.007
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606142101
Sample Date/Time: Tuesday, June 28, 2016 21:27:19
Number of Replicates: 3
Autosampler Position: 327
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86072.2	1.7				ug/L	81760	Standard
	Be	9	31.7	65.7	**0.0204**	0.019	95.2	ug/L	17	Standard
	Al	27	37149346.8	0.4	**228.9677**	4.521	2.0	ug/L	1410	Standard
	Sc	45	40693.0	3.6				ug/L	39543	Standard
	Ti	47	211.7	4.5	**0.7055**	0.031	4.4	ug/L	39	Standard
	V	51	19470.4	0.9	**2.6593**	0.027	1.0	ug/L	1415	Standard
	Cr	52	9577.4	0.5	**0.3506**	0.020	5.8	ug/L	6946	Standard
	Cr	53	13504.4	2.9	**15.3965**	0.370	2.4	ug/L	328	Standard
	Mn	55	2482687.2	1.9	**391.6677**	8.445	2.2	ug/L	823	Standard
	Co	59	1010.4	4.1	**0.1158**	0.006	5.6	ug/L	252	Standard
	Ni	60	2375.2	1.3	**1.1855**	0.029	2.4	ug/L	192	Standard
	Cu	65	1289.7	5.2	**0.6034**	0.030	5.0	ug/L	141	Standard
	Zn	66	3124.3	3.5	**2.0882**	0.083	4.0	ug/L	491	Standard
>	**Ge**	72	503476.5	0.9				ug/L	515087	Standard
	As	75	9860.1	0.9	**9.0129**	0.111	1.2	ug/L	20	Standard
	Se	82	502.3	2.2	**4.3447**	0.140	3.2	ug/L	15	Standard
	Se-1	77	853.4	2.0	**10.8689**	0.134	1.2	ug/L	48	Standard
>	**Ga**	71	115.0	32.8				mg/L	90	Standard
	Rb	85	655.0	2.3				ug/L	23	Standard
	Y	89	413248.4	1.0				ug/L	448510	Standard
>	**Rh**	103	123.3	18.3				ug/L	15	Standard
	Mo	98	77.7	17.1	**0.0125**	0.006	46.7	ug/L	11	Standard
	Ag	107	81.3	4.3	**-0.0029**	0.001	26.3	ug/L	134	Standard
	Cd	111	15.2	10.0	**0.0049**	0.001	14.6	mg/L	7	Standard
	Cd	114	38.3	10.8	**0.0032**	0.001	30.4	ug/L	16	Standard
>	**In**	115	557968.8	1.5				ug/L	611117	Standard
	Sn	118	745.0	12.3	**0.0359**	0.015	41.7	ug/L	518	Standard
	Sb	123	991.0	1.5	**0.2172**	0.002	1.1	ug/L	127	Standard
	Ba	135	220850.7	1.1	**117.1448**	0.521	0.4	ug/L	45	Standard
	Ce	140	1566.7	2.4				ug/L	298	Standard
>	**Tb**	159	920767.6	2.2				ug/L	977942	Standard
	Ho	165	58.3	38.7				ug/L	12	Standard
	Tl	203	55.7	23.2	**0.0023**	0.002	74.7	ug/L	31	Standard
	Tl	205	45.0	38.5	**0.0043**	0.003	60.7	ug/L	20	Standard
	Pb	206	169197.9	0.8	**33.4745**	0.521	1.6	ug/L	286	Standard
	Pb	207	142071.0	0.5	**31.1598**	0.218	0.7	ug/L	257	Standard
	Pb	208	579992.3	2.5	**32.8772**	0.549	1.7	ug/L	956	Standard
	U	238	648.7	2.7	**0.0940**	0.004	4.0	ug/L	17	Standard
>	**Bi**	209	520066.6	1.1				ug/L	557080	Standard

Sample ID: L1606142101
Report Date/Time: Tuesday, June 28, 2016 21:29:36
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.0185**	17.344	173.1	mg/L	0	Standard
Mg	24	83474.3	1.3	**108.0258**	4.532	4.2	mg/L	32	Standard
K	39	53.3	28.6	**0.0335**	0.050	147.9	mg/L	30	Standard
Ca	43	435.0	11.3	**58.7409**	4.908	8.4	mg/L	48	Standard
Fe	54	1115.5	8.0	**0.8464**	0.111	13.2	mg/L	159	Standard
Fe	57	721.7	4.0	**1.7451**	0.132	7.5	mg/L	195	Standard
Sc-1	45	40693.0	3.6				mg/L	39543	Standard
Cl	35	37196.3	1.1				ug/L	27642	Standard
Kr	83	3.0	66.7				ug/L	2	Standard
Br	81	117004.1	4.8				ug/L	1647	Standard
P	31	21849.1	2.8				ug/L	15231	Standard
S	34	2463.5	4.3				ug/L	1845	Standard
Sr	88	470.0	11.8				ug/L	70	Standard
C	12	400.0	4.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	47.0	20.8				mg/L	12	Standard
Ho-1	165	58.3	38.7				mg/L	12	Standard
Er	166	63.3	9.1				mg/L	20	Standard
I	127	290371.1	9.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		105.274	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		97.746	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606142101
Report Date/Time: Tuesday, June 28, 2016 21:29:36
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		91.303
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		93.356
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606142101
Report Date/Time: Tuesday, June 28, 2016 21:29:36
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 21:30:32
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	80064.5	3.2				ug/L	81760	Standard
	Be	9	49945.6	4.7	**51.3339**	1.054	2.1	ug/L	17	Standard
	Al	27	7446953.3	1.3	**49.3416**	1.021	2.1	ug/L	1410	Standard
	Sc	45	39994.4	3.1				ug/L	39543	Standard
	Ti	47	24563.6	1.8	**103.4611**	0.538	0.5	ug/L	39	Standard
	V	51	332459.7	1.4	**49.9760**	0.258	0.5	ug/L	1415	Standard
	Cr	52	346732.4	1.5	**51.2013**	0.219	0.4	ug/L	6946	Standard
	Cr	53	44092.4	3.1	**52.3486**	0.902	1.7	ug/L	328	Standard
	Mn	55	316773.7	1.0	**51.1225**	0.963	1.9	ug/L	823	Standard
	Co	59	338255.6	0.9	**50.4803**	0.452	0.9	ug/L	252	Standard
	Ni	60	91110.2	2.0	**50.6466**	0.186	0.4	ug/L	192	Standard
	Cu	65	91902.3	2.1	**50.6488**	0.160	0.3	ug/L	141	Standard
	Zn	66	53757.9	1.1	**50.6325**	0.663	1.3	ug/L	491	Standard
>	Ge	72	490984.9	1.8				ug/L	515087	Standard
	As	75	53512.8	1.1	**49.9767**	0.363	0.7	ug/L	20	Standard
	Se	82	5576.6	0.4	**50.7061**	0.891	1.8	ug/L	15	Standard
	Se-1	77	3693.5	2.8	**50.7124**	1.021	2.0	ug/L	48	Standard
>	Ga	71	63.3	29.9				mg/L	90	Standard
	Rb	85	831.7	5.0				ug/L	23	Standard
	Y	89	421530.5	1.2				ug/L	448510	Standard
>	Rh	103	28.3	44.4				ug/L	15	Standard
	Mo	98	254944.7	0.2	**97.5527**	0.365	0.4	ug/L	11	Standard
	Ag	107	312741.7	1.2	**48.1812**	0.687	1.4	ug/L	134	Standard
	Cd	111	100992.4	1.1	**48.8720**	0.473	1.0	mg/L	7	Standard
	Cd	114	246539.5	0.8	**49.0819**	0.317	0.6	ug/L	16	Standard
>	In	115	590755.4	0.3				ug/L	611117	Standard
	Sn	118	285297.8	0.9	**49.8102**	0.309	0.6	ug/L	518	Standard
	Sb	123	223087.7	1.5	**48.8998**	0.798	1.6	ug/L	127	Standard
	Ba	135	97286.0	1.5	**48.7240**	0.868	1.8	ug/L	45	Standard
	Ce	140	78.3	36.9				ug/L	298	Standard
>	Tb	159	940662.6	1.7				ug/L	977942	Standard
	Ho	165	20.0	25.0				ug/L	12	Standard
	Tl	203	402817.7	0.9	**49.5894**	0.553	1.1	ug/L	31	Standard
	Tl	205	361515.5	0.7	**51.2534**	0.688	1.3	ug/L	20	Standard
	Pb	206	256836.7	1.0	**49.2334**	0.527	1.1	ug/L	286	Standard
	Pb	207	233772.2	1.4	**49.6819**	0.304	0.6	ug/L	257	Standard
	Pb	208	939106.3	0.8	**51.5898**	0.228	0.4	ug/L	956	Standard
	U	238	351652.3	0.3	**50.9098**	0.427	0.8	ug/L	17	Standard
>	Bi	209	537054.5	1.2				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:32:49
Page 1

Approved: June 29, 2016

Na	23	5.0	173.2	**32.8471**	56.884	173.2	mg/L	0	Standard
Mg	24	4075.5	3.9	**5.3128**	0.340	6.4	mg/L	32	Standard
K	39	1373.4	2.2	**4.4475**	0.053	1.2	mg/L	30	Standard
Ca	43	71.7	22.4	**5.4062**	2.324	43.0	mg/L	48	Standard
Fe	54	5589.4	3.1	**4.8672**	0.205	4.2	mg/L	159	Standard
Fe	57	1635.1	4.9	**4.8331**	0.203	4.2	mg/L	195	Standard
> **Sc-1**	45	39994.4	3.1				mg/L	39543	Standard
Cl	35	35754.8	4.0				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	3610.5	34.2				ug/L	1647	Standard
P	31	19324.0	4.2				ug/L	15231	Standard
S	34	2706.9	5.1				ug/L	1845	Standard
Sr	88	63.3	27.7				ug/L	70	Standard
C	12	143.3	42.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	5.9	100.8				mg/L	12	Standard
Ho-1	165	20.0	25.0				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	9711.8	44.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	102.668		
Al	27	98.683		
Sc	45			
Ti	47	103.461		
V	51	99.952		
Cr	52	102.403		
Cr	53			
Mn	55	102.245		
Co	59	100.961		
Ni	60	101.293		
Cu	65	101.298		
Zn	66	101.265		
> Ge	72		95.321	
As	75	99.953		
Se	82	101.412		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:32:49
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	97.553		
Ag	107	96.362		
Cd	111	97.744		
Cd	114			
> In	115			96.668
Sn	118	99.620		
Sb	123	97.800		
Ba	135	97.448		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	99.179		
Tl	205			
Pb	206			
Pb	207			
Pb	208	103.180		
U	238	101.820		
> Bi	209			96.405
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:32:49
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 21:33:43
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77306.8	2.0				ug/L	81760	Standard
	Be	9	21.7	26.6	**0.0134**	0.006	48.3	ug/L	17	Standard
	Al	27	7303.5	27.4	**0.0355**	0.013	36.3	ug/L	1410	Standard
	Sc	45	38198.1	0.8				ug/L	39543	Standard
	Ti	47	25.7	25.3	**-0.0566**	0.027	47.6	ug/L	39	Standard
	V	51	1109.0	7.8	**-0.0386**	0.015	38.5	ug/L	1415	Standard
	Cr	52	6097.9	0.7	**-0.1350**	0.015	10.8	ug/L	6946	Standard
	Cr	53	763.4	6.1	**0.5733**	0.039	6.9	ug/L	328	Standard
	Mn	55	836.4	10.1	**-0.0128**	0.011	87.4	ug/L	823	Standard
	Co	59	230.3	21.5	**0.0031**	0.007	222.1	ug/L	252	Standard
	Ni	60	176.3	3.3	**-0.0064**	0.005	75.2	ug/L	192	Standard
	Cu	65	147.3	9.8	**-0.0093**	0.007	71.2	ug/L	141	Standard
	Zn	66	414.7	6.5	**-0.4297**	0.019	4.3	ug/L	491	Standard
>	**Ge**	72	489415.0	1.8				ug/L	515087	Standard
	As	75	-13.4	315.6	**0.0274**	0.040	146.9	ug/L	20	Standard
	Se	82	21.4	39.0	**0.0756**	0.076	100.0	ug/L	15	Standard
	Se-1	77	62.3	3.3	**0.1522**	0.035	23.1	ug/L	48	Standard
>	**Ga**	71	55.0	41.7				mg/L	90	Standard
	Rb	85	21.7	26.6				ug/L	23	Standard
	Y	89	425437.4	1.2				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	273.7	29.6	**0.0854**	0.031	36.3	ug/L	11	Standard
	Ag	107	107.7	30.0	**0.0003**	0.005	1550.1	ug/L	134	Standard
	Cd	111	12.2	52.2	**0.0030**	0.003	103.1	mg/L	7	Standard
	Cd	114	23.6	69.2	**-0.0002**	0.003	1385.1	ug/L	16	Standard
>	**In**	115	592746.5	0.3				ug/L	611117	Standard
	Sn	118	735.0	15.8	**0.0262**	0.020	76.7	ug/L	518	Standard
	Sb	123	418.9	32.2	**0.0787**	0.029	37.2	ug/L	127	Standard
	Ba	135	58.7	27.6	**0.0042**	0.008	193.4	ug/L	45	Standard
	Ce	140	16.7	34.6				ug/L	298	Standard
>	**Tb**	159	928666.7	1.6				ug/L	977942	Standard
	Ho	165	11.7	89.2				ug/L	12	Standard
	Tl	203	77.0	34.1	**0.0046**	0.003	68.4	ug/L	31	Standard
	Tl	205	76.7	35.9	**0.0085**	0.004	45.8	ug/L	20	Standard
	Pb	206	287.0	11.8	**-0.0088**	0.007	78.2	ug/L	286	Standard
	Pb	207	252.0	6.9	**-0.0047**	0.004	83.8	ug/L	257	Standard
	Pb	208	1038.7	6.8	**-0.0037**	0.004	102.1	ug/L	956	Standard
	U	238	52.3	19.6	**0.0046**	0.001	32.5	ug/L	17	Standard
>	**Bi**	209	540535.6	1.1				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:36:00
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**11.1254**	19.261	173.1	mg/L	0	Standard
Mg	24	41.7	25.0	**0.0002**	0.014	5858.2	mg/L	32	Standard
K	39	25.0	20.0	**-0.0539**	0.018	33.4	mg/L	30	Standard
Ca	43	40.0	25.0	**0.9508**	1.613	169.6	mg/L	48	Standard
Fe	54	117.8	25.5	**-0.0268**	0.028	102.8	mg/L	159	Standard
Fe	57	235.0	7.7	**0.1964**	0.057	28.8	mg/L	195	Standard
> **Sc-1**	45	38198.1	0.8				mg/L	39543	Standard
Cl	35	35245.6	1.5				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	2320.2	7.7				ug/L	1647	Standard
P	31	18950.2	2.2				ug/L	15231	Standard
S	34	2475.2	4.2				ug/L	1845	Standard
Sr	88	68.3	27.7				ug/L	70	Standard
C	12	103.3	5.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	0	Standard
Dy	164	8.9	118.0				mg/L	12	Standard
Ho-1	165	11.7	89.2				mg/L	12	Standard
Er	166	23.3	49.5				mg/L	20	Standard
I	127	4469.0	13.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.016	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:36:00
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.994
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.030
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:36:00
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: PBW 82 WG574205-02
Sample Date/Time: Tuesday, June 28, 2016 21:36:55
Number of Replicates: 3
Autosampler Position: 328
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77750.7	1.6				ug/L	81760	Standard
	Be	9	23.3	24.7	**0.0149**	0.006	38.5	ug/L	17	Standard
	Al	27	11095.9	27.0	**0.0614**	0.021	33.8	ug/L	1410	Standard
	Sc	45	39309.3	3.2				ug/L	39543	Standard
	Ti	47	35.7	19.9	**-0.0160**	0.027	171.9	ug/L	39	Standard
	V	51	1269.4	3.8	**-0.0162**	0.012	71.1	ug/L	1415	Standard
	Cr	52	7452.2	2.1	**0.0582**	0.015	26.3	ug/L	6946	Standard
	Cr	53	805.0	4.3	**0.6140**	0.043	7.0	ug/L	328	Standard
	Mn	55	3268.0	4.0	**0.3772**	0.034	9.1	ug/L	823	Standard
	Co	59	296.7	4.7	**0.0126**	0.001	9.3	ug/L	252	Standard
	Ni	60	300.3	5.8	**0.0613**	0.013	21.5	ug/L	192	Standard
	Cu	65	252.7	4.2	**0.0478**	0.009	19.3	ug/L	141	Standard
	Zn	66	1868.1	1.9	**0.9480**	0.028	3.0	ug/L	491	Standard
>	**Ge**	72	494530.1	2.9				ug/L	515087	Standard
	As	75	-37.6	66.7	**0.0055**	0.023	429.2	ug/L	20	Standard
	Se	82	15.9	27.5	**0.0238**	0.038	158.5	ug/L	15	Standard
	Se-1	77	56.7	8.0	**0.0642**	0.051	79.2	ug/L	48	Standard
>	**Ga**	71	51.7	31.1				mg/L	90	Standard
	Rb	85	45.0	48.4				ug/L	23	Standard
	Y	89	413936.1	2.9				ug/L	448510	Standard
>	**Rh**	103	23.3	32.7				ug/L	15	Standard
	Mo	98	83.2	8.2	**0.0127**	0.003	23.0	ug/L	11	Standard
	Ag	107	95.3	10.0	**-0.0016**	0.002	102.9	ug/L	134	Standard
	Cd	111	39.5	14.4	**0.0162**	0.003	18.5	mg/L	7	Standard
	Cd	114	83.7	13.3	**0.0117**	0.002	21.2	ug/L	16	Standard
>	**In**	115	592344.0	1.8				ug/L	611117	Standard
	Sn	118	795.0	11.2	**0.0369**	0.017	47.0	ug/L	518	Standard
	Sb	123	176.9	20.5	**0.0258**	0.007	28.3	ug/L	127	Standard
	Ba	135	88.3	8.2	**0.0191**	0.004	22.5	ug/L	45	Standard
	Ce	140	48.3	15.8				ug/L	298	Standard
>	**Tb**	159	936174.2	1.5				ug/L	977942	Standard
	Ho	165	6.7	114.6				ug/L	12	Standard
	Tl	203	128.0	4.9	**0.0106**	0.001	7.1	ug/L	31	Standard
	Tl	205	116.7	44.6	**0.0139**	0.007	52.1	ug/L	20	Standard
	Pb	206	358.0	3.9	**0.0036**	0.003	73.1	ug/L	286	Standard
	Pb	207	341.0	6.5	**0.0129**	0.005	36.4	ug/L	257	Standard
	Pb	208	1292.4	1.6	**0.0090**	0.001	13.2	ug/L	956	Standard
	U	238	93.7	24.0	**0.0103**	0.003	31.1	ug/L	17	Standard
>	**Bi**	209	549438.7	0.1				ug/L	557080	Standard

Sample ID: PBW 82 WG574205-02
Report Date/Time: Tuesday, June 28, 2016 21:39:12
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.5912**	18.336	173.1	mg/L	0	Standard
Mg	24	55.0	47.2	**0.0170**	0.036	211.2	mg/L	32	Standard
K	39	38.3	52.7	**-0.0104**	0.070	678.8	mg/L	30	Standard
Ca	43	46.7	16.4	**1.7858**	1.139	63.8	mg/L	48	Standard
Fe	54	125.7	11.9	**-0.0224**	0.015	66.3	mg/L	159	Standard
Fe	57	226.7	12.1	**0.1455**	0.093	63.6	mg/L	195	Standard
> **Sc-1**	45	39309.3	3.2				mg/L	39543	Standard
Cl	35	34398.3	0.8				ug/L	27642	Standard
Kr	83	3.0	66.7				ug/L	2	Standard
Br	81	1940.1	5.9				ug/L	1647	Standard
P	31	18074.1	0.8				ug/L	15231	Standard
S	34	2498.5	3.0				ug/L	1845	Standard
Sr	88	61.7	46.1				ug/L	70	Standard
C	12	113.3	13.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	9.2	3.0				mg/L	12	Standard
Ho-1	165	6.7	114.6				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	3242.0	12.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		95.096	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.009	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBW 82 WG574205-02
Report Date/Time: Tuesday, June 28, 2016 21:39:12
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.928
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.628
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 82 WG574205-02
Report Date/Time: Tuesday, June 28, 2016 21:39:12
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: LCSW 82 WG574205-03
Sample Date/Time: Tuesday, June 28, 2016 21:40:07
Number of Replicates: 3
Autosampler Position: 329
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77370.5	2.5				ug/L	81760	Standard
	Be	9	47069.7	3.6	**50.0685**	0.762	1.5	ug/L	17	Standard
	Al	27	8339.0	0.8	**0.0428**	0.001	2.3	ug/L	1410	Standard
	Sc	45	38442.1	2.3				ug/L	39543	Standard
	Ti	47	32.3	15.3	**-0.0274**	0.023	83.9	ug/L	39	Standard
	V	51	328373.9	1.3	**49.7728**	1.098	2.2	ug/L	1415	Standard
	Cr	52	340284.0	1.1	**50.6522**	0.528	1.0	ug/L	6946	Standard
	Cr	53	43268.4	1.2	**51.8066**	1.337	2.6	ug/L	328	Standard
	Mn	55	314954.5	0.2	**51.2439**	0.720	1.4	ug/L	823	Standard
	Co	59	331722.8	0.6	**49.9116**	0.628	1.3	ug/L	252	Standard
	Ni	60	91188.9	0.7	**51.1120**	0.365	0.7	ug/L	192	Standard
	Cu	65	91760.4	1.1	**50.9900**	0.227	0.4	ug/L	141	Standard
	Zn	66	54150.6	1.3	**51.4287**	0.106	0.2	ug/L	491	Standard
>	**Ge**	72	486985.3	1.4				ug/L	515087	Standard
	As	75	52577.5	0.5	**49.5070**	0.430	0.9	ug/L	20	Standard
	Se	82	5439.8	1.4	**49.8572**	0.058	0.1	ug/L	15	Standard
	Se-1	77	3674.8	1.0	**50.8806**	0.838	1.6	ug/L	48	Standard
>	**Ga**	71	70.0	21.4				mg/L	90	Standard
	Rb	85	35.0	37.8				ug/L	23	Standard
	Y	89	431830.9	0.7				ug/L	448510	Standard
>	**Rh**	103	38.3	15.1				ug/L	15	Standard
	Mo	98	69.3	10.5	**0.0074**	0.003	42.1	ug/L	11	Standard
	Ag	107	312423.9	0.2	**48.0443**	1.285	2.7	ug/L	134	Standard
	Cd	111	99250.2	0.2	**47.9394**	1.102	2.3	mg/L	7	Standard
	Cd	114	241436.3	0.5	**47.9735**	0.947	2.0	ug/L	16	Standard
>	**In**	115	592091.2	2.5				ug/L	611117	Standard
	Sn	118	838.4	9.3	**0.0444**	0.014	30.4	ug/L	518	Standard
	Sb	123	213126.5	0.4	**46.6279**	1.103	2.4	ug/L	127	Standard
	Ba	135	94941.2	1.6	**47.4531**	1.010	2.1	ug/L	45	Standard
	Ce	140	68.3	22.4				ug/L	298	Standard
>	**Tb**	159	920874.0	1.0				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	395784.4	0.5	**49.2144**	0.506	1.0	ug/L	31	Standard
	Tl	205	352148.8	0.9	**50.4246**	0.261	0.5	ug/L	20	Standard
	Pb	206	261292.0	1.2	**50.5931**	0.633	1.3	ug/L	286	Standard
	Pb	207	229140.8	1.1	**49.1895**	0.516	1.0	ug/L	257	Standard
	Pb	208	939702.2	0.3	**52.1447**	0.484	0.9	ug/L	956	Standard
	U	238	338970.6	0.7	**49.5659**	0.166	0.3	ug/L	17	Standard
>	**Bi**	209	531689.3	0.7				ug/L	557080	Standard

Sample ID: LCSW 82 WG574205-03
Report Date/Time: Tuesday, June 28, 2016 21:42:24
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	35.0	0.0	**-0.0092**	0.001	12.0	mg/L	32	Standard
K	39	35.0	42.9	**-0.0195**	0.052	268.3	mg/L	30	Standard
Ca	43	38.3	32.8	**0.6220**	1.885	303.1	mg/L	48	Standard
Fe	54	132.7	8.8	**-0.0133**	0.014	103.8	mg/L	159	Standard
Fe	57	226.7	10.4	**0.1628**	0.084	51.6	mg/L	195	Standard
Sc-1	45	38442.1	2.3				mg/L	39543	Standard
Cl	35	35188.8	3.0				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1823.4	6.0				ug/L	1647	Standard
P	31	17611.9	0.6				ug/L	15231	Standard
S	34	2310.2	2.3				ug/L	1845	Standard
Sr	88	75.0	17.6				ug/L	70	Standard
C	12	96.7	46.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	5.9	194.5				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	2716.9	12.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.631	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.544	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 82 WG574205-03
Report Date/Time: Tuesday, June 28, 2016 21:42:24
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.887
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.442
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 82 WG574205-03
Report Date/Time: Tuesday, June 28, 2016 21:42:24
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606134301 WG574205-01
Sample Date/Time: Tuesday, June 28, 2016 21:43:18
Number of Replicates: 3
Autosampler Position: 330
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74323.5	0.8				ug/L	81760	Standard
	Be	9	31.7	36.5	**0.0253**	0.013	49.8	ug/L	17	Standard
	Al	27	2998278.2	0.6	**21.3833**	0.050	0.2	ug/L	1410	Standard
	Sc	45	38510.6	3.3				ug/L	39543	Standard
	Ti	47	232.7	4.1	**0.8126**	0.032	3.9	ug/L	39	Standard
	V	51	1652.5	23.4	**0.0414**	0.055	131.6	ug/L	1415	Standard
	Cr	52	135321.3	0.7	**19.2638**	0.330	1.7	ug/L	6946	Standard
	Cr	53	19187.1	1.5	**22.5044**	0.281	1.2	ug/L	328	Standard
	Mn	55	3959.9	4.4	**0.4904**	0.040	8.1	ug/L	823	Standard
	Co	59	833.7	4.3	**0.0927**	0.008	8.3	ug/L	252	Standard
	Ni	60	44655.7	0.7	**24.6837**	0.632	2.6	ug/L	192	Standard
	Cu	65	31746.9	0.7	**17.3740**	0.276	1.6	ug/L	141	Standard
	Zn	66	57137.2	0.4	**53.6669**	1.121	2.1	ug/L	491	Standard
>	**Ge**	72	492860.3	1.9				ug/L	515087	Standard
	As	75	122.8	20.0	**0.1545**	0.023	14.7	ug/L	20	Standard
	Se	82	36.2	8.8	**0.2097**	0.035	16.8	ug/L	15	Standard
	Se-1	77	205.3	7.5	**2.1323**	0.253	11.9	ug/L	48	Standard
>	**Ga**	71	501.7	15.2				mg/L	90	Standard
	Rb	85	10964.0	0.9				ug/L	23	Standard
	Y	89	418533.1	1.8				ug/L	448510	Standard
>	**Rh**	103	56.7	31.0				ug/L	15	Standard
	Mo	98	2829.4	1.2	**1.1053**	0.009	0.8	ug/L	11	Standard
	Ag	107	25274.7	1.2	**4.0283**	0.033	0.8	ug/L	134	Standard
	Cd	111	119.1	9.1	**0.0569**	0.005	9.6	mg/L	7	Standard
	Cd	114	490.9	6.4	**0.0966**	0.006	6.7	ug/L	16	Standard
>	**In**	115	568910.2	0.4				ug/L	611117	Standard
	Sn	118	49674.7	2.3	**8.9227**	0.228	2.6	ug/L	518	Standard
	Sb	123	2415.2	17.6	**0.5368**	0.095	17.6	ug/L	127	Standard
	Ba	135	20410.1	0.9	**10.5946**	0.091	0.9	ug/L	45	Standard
	Ce	140	566.7	11.5				ug/L	298	Standard
>	**Tb**	159	895351.1	0.7				ug/L	977942	Standard
	Ho	165	20.0	25.0				ug/L	12	Standard
	Tl	203	216.3	14.7	**0.0231**	0.004	18.6	ug/L	31	Standard
	Tl	205	211.7	17.7	**0.0291**	0.005	18.1	ug/L	20	Standard
	Pb	206	14582.7	1.0	**2.8682**	0.053	1.8	ug/L	286	Standard
	Pb	207	11703.2	0.8	**2.5502**	0.014	0.5	ug/L	257	Standard
	Pb	208	49274.0	0.6	**2.7782**	0.042	1.5	ug/L	956	Standard
	U	238	116.0	23.9	**0.0146**	0.004	29.7	ug/L	17	Standard
>	**Bi**	209	512746.8	0.9				ug/L	557080	Standard

Sample ID: L1606134301 WG574205-01
Report Date/Time: Tuesday, June 28, 2016 21:45:35
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	23663.5	1.4	**32.3162**	1.385	4.3	mg/L	32	Standard
K	39	683.3	9.3	**2.2286**	0.197	8.8	mg/L	30	Standard
Ca	43	95.0	32.9	**9.3774**	4.330	46.2	mg/L	48	Standard
Fe	54	183.2	7.6	**0.0331**	0.009	27.1	mg/L	159	Standard
Fe	57	263.3	14.4	**0.2873**	0.119	41.4	mg/L	195	Standard
> **Sc-1**	45	38510.6	3.3				mg/L	39543	Standard
Cl	35	35170.8	1.9				ug/L	27642	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	3347.0	4.4				ug/L	1647	Standard
P	31	18034.1	2.9				ug/L	15231	Standard
S	34	2231.8	5.1				ug/L	1845	Standard
Sr	88	143.3	22.7				ug/L	70	Standard
C	12	123.3	47.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	12.4	90.1				mg/L	12	Standard
Ho-1	165	20.0	25.0				mg/L	12	Standard
Er	166	20.0	86.6				mg/L	20	Standard
I	127	33806.0	2.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		90.904	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.685	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606134301 WG574205-01
Report Date/Time: Tuesday, June 28, 2016 21:45:35
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		93.094
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		92.042
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606134301DP WG574205-04
Sample Date/Time: Tuesday, June 28, 2016 21:46:30
Number of Replicates: 3
Autosampler Position: 331
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75650.1	1.3				ug/L	81760	Standard
	Be	9	38.3	7.5	**0.0320**	0.004	11.5	ug/L	17	Standard
	Al	27	3034197.2	1.7	**21.2640**	0.568	2.7	ug/L	1410	Standard
	Sc	45	38358.5	1.6				ug/L	39543	Standard
	Ti	47	242.3	3.4	**0.8481**	0.030	3.6	ug/L	39	Standard
	V	51	2921.5	21.4	**0.2301**	0.091	39.7	ug/L	1415	Standard
	Cr	52	132102.7	0.7	**18.6755**	0.034	0.2	ug/L	6946	Standard
	Cr	53	19672.8	2.2	**22.9627**	0.596	2.6	ug/L	328	Standard
	Mn	55	13082.6	2.2	**1.9497**	0.036	1.8	ug/L	823	Standard
	Co	59	922.0	4.8	**0.1051**	0.006	5.6	ug/L	252	Standard
	Ni	60	44071.0	1.0	**24.2264**	0.187	0.8	ug/L	192	Standard
	Cu	65	31222.8	1.1	**16.9947**	0.096	0.6	ug/L	141	Standard
	Zn	66	55768.5	0.9	**52.0764**	0.197	0.4	ug/L	491	Standard
>	**Ge**	72	495382.5	0.5				ug/L	515087	Standard
	As	75	105.3	17.2	**0.1377**	0.017	12.5	ug/L	20	Standard
	Se	82	31.4	9.3	**0.1642**	0.025	15.2	ug/L	15	Standard
	Se-1	77	224.7	3.6	**2.3824**	0.126	5.3	ug/L	48	Standard
>	**Ga**	71	485.0	9.9				mg/L	90	Standard
	Rb	85	10640.4	0.2				ug/L	23	Standard
	Y	89	407749.6	1.6				ug/L	448510	Standard
>	**Rh**	103	56.7	27.0				ug/L	15	Standard
	Mo	98	2759.9	2.4	**1.0902**	0.013	1.1	ug/L	11	Standard
	Ag	107	24718.5	0.6	**3.9851**	0.067	1.7	ug/L	134	Standard
	Cd	111	119.1	7.2	**0.0577**	0.005	8.8	mg/L	7	Standard
	Cd	114	389.6	30.3	**0.0764**	0.024	30.9	ug/L	16	Standard
>	**In**	115	562489.2	1.2				ug/L	611117	Standard
	Sn	118	48104.7	5.2	**8.7400**	0.532	6.1	ug/L	518	Standard
	Sb	123	1085.9	11.8	**0.2370**	0.026	11.2	ug/L	127	Standard
	Ba	135	21279.9	0.9	**11.1745**	0.156	1.4	ug/L	45	Standard
	Ce	140	1446.7	1.7				ug/L	298	Standard
>	**Tb**	159	899809.7	0.9				ug/L	977942	Standard
	Ho	165	38.3	37.7				ug/L	12	Standard
	Tl	203	211.3	8.9	**0.0223**	0.003	11.4	ug/L	31	Standard
	Tl	205	171.7	14.4	**0.0230**	0.004	16.6	ug/L	20	Standard
	Pb	206	14329.4	1.7	**2.7934**	0.069	2.5	ug/L	286	Standard
	Pb	207	11471.7	1.5	**2.4775**	0.050	2.0	ug/L	257	Standard
	Pb	208	48196.5	1.3	**2.6929**	0.035	1.3	ug/L	956	Standard
	U	238	162.0	9.3	**0.0214**	0.002	11.2	ug/L	17	Standard
>	**Bi**	209	517046.0	0.7				ug/L	557080	Standard

Sample ID: L1606134301DP WG574205-04
Report Date/Time: Tuesday, June 28, 2016 21:48:47
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	23364.7	1.2	**32.0144**	0.886	2.8	mg/L	32	Standard
K	39	620.0	6.9	**2.0197**	0.176	8.7	mg/L	30	Standard
Ca	43	98.3	39.5	**10.0220**	6.022	60.1	mg/L	48	Standard
Fe	54	175.4	14.5	**0.0269**	0.026	96.7	mg/L	159	Standard
Fe	57	298.3	7.6	**0.4146**	0.094	22.7	mg/L	195	Standard
> **Sc-1**	45	38358.5	1.6				mg/L	39543	Standard
Cl	35	35843.0	1.2				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	4177.2	4.7				ug/L	1647	Standard
P	31	18070.8	1.2				ug/L	15231	Standard
S	34	2273.5	5.8				ug/L	1845	Standard
Sr	88	125.0	20.0				ug/L	70	Standard
C	12	143.3	21.3				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	16.7	91.7				mg/L	0	Standard
Dy	164	45.4	50.3				mg/L	12	Standard
Ho-1	165	38.3	37.7				mg/L	12	Standard
Er	166	26.7	21.7				mg/L	20	Standard
I	127	32309.4	0.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		92.527	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.175	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606134301DP WG574205-04
Report Date/Time: Tuesday, June 28, 2016 21:48:47
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		92.043
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		92.814
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606134301S WG574205-05
Sample Date/Time: Tuesday, June 28, 2016 21:49:41
Number of Replicates: 3
Autosampler Position: 332
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	76159.4	2.8				ug/L	81760	Standard
	Be	9	48450.7	1.7	**52.3785**	0.949	1.8	ug/L	17	Standard
	Al	27	2895043.9	1.0	**20.1618**	0.737	3.7	ug/L	1410	Standard
	Sc	45	38742.8	1.9				ug/L	39543	Standard
	Ti	47	240.0	7.2	**0.8720**	0.082	9.4	ug/L	39	Standard
	V	51	323841.2	0.8	**49.8512**	0.902	1.8	ug/L	1415	Standard
	Cr	52	453207.8	1.4	**68.8842**	0.591	0.9	ug/L	6946	Standard
	Cr	53	60472.1	1.8	**73.6607**	0.780	1.1	ug/L	328	Standard
	Mn	55	309167.0	1.2	**51.0846**	0.705	1.4	ug/L	823	Standard
	Co	59	322466.2	0.8	**49.2767**	0.754	1.5	ug/L	252	Standard
	Ni	60	129806.7	1.0	**73.9391**	0.553	0.7	ug/L	192	Standard
	Cu	65	118360.2	1.2	**66.8253**	0.317	0.5	ug/L	141	Standard
	Zn	66	105137.5	0.6	**102.2235**	1.030	1.0	ug/L	491	Standard
>	Ge	72	479488.8	1.1				ug/L	515087	Standard
	As	75	53941.4	1.7	**51.5891**	1.400	2.7	ug/L	20	Standard
	Se	82	5350.1	0.9	**49.8082**	0.982	2.0	ug/L	15	Standard
	Se-1	77	3727.1	2.1	**52.4361**	1.513	2.9	ug/L	48	Standard
>	Ga	71	476.7	2.2				mg/L	90	Standard
	Rb	85	10533.7	3.9				ug/L	23	Standard
	Y	89	417307.7	1.3				ug/L	448510	Standard
>	Rh	103	73.3	25.8				ug/L	15	Standard
	Mo	98	2762.6	0.6	**1.0591**	0.012	1.2	ug/L	11	Standard
	Ag	107	321242.3	0.9	**50.4672**	0.741	1.5	ug/L	134	Standard
	Cd	111	96845.5	0.4	**47.7898**	0.498	1.0	mg/L	7	Standard
	Cd	114	233334.7	1.1	**47.3688**	0.646	1.4	ug/L	16	Standard
>	In	115	579360.0	0.8				ug/L	611117	Standard
	Sn	118	55146.4	13.8	**9.7304**	1.292	13.3	ug/L	518	Standard
	Sb	123	221114.9	0.1	**49.4220**	0.385	0.8	ug/L	127	Standard
	Ba	135	112895.7	0.5	**57.6594**	0.626	1.1	ug/L	45	Standard
	Ce	140	313.3	11.5				ug/L	298	Standard
>	Tb	159	912553.5	0.8				ug/L	977942	Standard
	Ho	165	15.0	66.7				ug/L	12	Standard
	Tl	203	384933.8	0.5	**49.0255**	0.608	1.2	ug/L	31	Standard
	Tl	205	344367.9	0.4	**50.5097**	0.783	1.5	ug/L	20	Standard
	Pb	206	266968.5	0.3	**52.9500**	0.690	1.3	ug/L	286	Standard
	Pb	207	232893.3	0.6	**51.2096**	0.463	0.9	ug/L	257	Standard
	Pb	208	948808.9	1.4	**53.9303**	1.097	2.0	ug/L	956	Standard
	U	238	335133.4	0.9	**50.1950**	0.747	1.5	ug/L	17	Standard
>	Bi	209	519140.8	1.4				ug/L	557080	Standard

Sample ID: L1606134301S WG574205-05
Report Date/Time: Tuesday, June 28, 2016 21:51:58
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	23122.6	1.5	**31.3670**	0.821	2.6	mg/L	32	Standard
K	39	550.0	13.4	**1.7528**	0.224	12.8	mg/L	30	Standard
Ca	43	93.3	3.1	**9.1209**	0.222	2.4	mg/L	48	Standard
Fe	54	147.8	38.5	**0.0002**	0.055	34709.4	mg/L	159	Standard
Fe	57	261.7	11.7	**0.2782**	0.120	43.1	mg/L	195	Standard
> **Sc-1**	45	38742.8	1.9				mg/L	39543	Standard
Cl	35	35887.1	0.6				ug/L	27642	Standard
Kr	83	2.3	89.2				ug/L	2	Standard
Br	81	3013.6	4.3				ug/L	1647	Standard
P	31	18511.3	3.1				ug/L	15231	Standard
S	34	2173.5	7.9				ug/L	1845	Standard
Sr	88	115.0	4.3				ug/L	70	Standard
C	12	143.3	35.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	12.7	50.0				mg/L	12	Standard
Ho-1	165	15.0	66.7				mg/L	12	Standard
Er	166	13.3	114.6				mg/L	20	Standard
I	127	29318.3	1.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.150	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.089	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606134301S WG574205-05
Report Date/Time: Tuesday, June 28, 2016 21:51:58
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.804
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		93.190
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606134301S WG574205-05
Report Date/Time: Tuesday, June 28, 2016 21:51:58
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606134301SD WG574205-06
Sample Date/Time: Tuesday, June 28, 2016 21:52:53
Number of Replicates: 3
Autosampler Position: 333
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74233.3	4.4				ug/L	81760	Standard
	Be	9	46944.3	0.7	**52.1161**	2.208	4.2	ug/L	17	Standard
	Al	27	2867977.1	1.8	**20.5130**	1.211	5.9	ug/L	1410	Standard
	Sc	45	37107.1	4.7				ug/L	39543	Standard
	Ti	47	208.3	2.4	**0.7439**	0.023	3.1	ug/L	39	Standard
	V	51	314357.5	0.7	**48.8632**	0.269	0.5	ug/L	1415	Standard
	Cr	52	443013.9	1.2	**67.9920**	0.699	1.0	ug/L	6946	Standard
	Cr	53	57726.4	1.5	**71.0049**	0.771	1.1	ug/L	328	Standard
	Mn	55	299003.6	1.5	**49.8899**	0.541	1.1	ug/L	823	Standard
	Co	59	310607.5	1.9	**47.9296**	0.658	1.4	ug/L	252	Standard
	Ni	60	125816.0	1.4	**72.3742**	0.740	1.0	ug/L	192	Standard
	Cu	65	116930.9	2.0	**66.6728**	1.005	1.5	ug/L	141	Standard
	Zn	66	105389.9	0.9	**103.4925**	0.359	0.3	ug/L	491	Standard
>	**Ge**	72	474757.8	0.5				ug/L	515087	Standard
	As	75	52320.0	1.0	**50.5276**	0.214	0.4	ug/L	20	Standard
	Se	82	5199.8	1.3	**48.8839**	0.776	1.6	ug/L	15	Standard
	Se-1	77	3645.4	2.7	**51.7794**	1.336	2.6	ug/L	48	Standard
>	**Ga**	71	471.7	1.2				mg/L	90	Standard
	Rb	85	10923.9	0.7				ug/L	23	Standard
	Y	89	414912.1	1.8				ug/L	448510	Standard
>	**Rh**	103	51.7	22.3				ug/L	15	Standard
	Mo	98	2810.6	3.1	**1.1098**	0.050	4.5	ug/L	11	Standard
	Ag	107	309794.3	1.1	**50.0964**	1.925	3.8	ug/L	134	Standard
	Cd	111	92939.1	0.1	**47.2004**	1.405	3.0	mg/L	7	Standard
	Cd	114	227712.2	0.8	**47.5835**	1.757	3.7	ug/L	16	Standard
>	**In**	115	563239.5	3.0				ug/L	611117	Standard
	Sn	118	48800.3	6.0	**8.8657**	0.749	8.5	ug/L	518	Standard
	Sb	123	213681.9	0.5	**49.1574**	1.598	3.3	ug/L	127	Standard
	Ba	135	108173.8	0.8	**56.8630**	1.899	3.3	ug/L	45	Standard
	Ce	140	258.3	11.8				ug/L	298	Standard
>	**Tb**	159	874543.1	3.3				ug/L	977942	Standard
	Ho	165	16.7	62.4				ug/L	12	Standard
	Tl	203	370949.6	0.9	**49.2226**	1.185	2.4	ug/L	31	Standard
	Tl	205	332170.0	0.6	**50.7579**	1.084	2.1	ug/L	20	Standard
	Pb	206	260124.0	0.8	**53.7518**	1.191	2.2	ug/L	286	Standard
	Pb	207	226333.8	1.1	**51.8538**	1.336	2.6	ug/L	257	Standard
	Pb	208	924176.8	1.2	**54.7310**	1.509	2.8	ug/L	956	Standard
	U	238	327055.2	1.2	**51.0349**	1.235	2.4	ug/L	17	Standard
>	**Bi**	209	498348.0	1.6				ug/L	557080	Standard

Sample ID: L1606134301SD WG574205-06
Report Date/Time: Tuesday, June 28, 2016 21:55:09
Page 1

Approved: June 29, 2016

Na	23	5.0	173.2	**33.3751**	57.799	173.2	mg/L	0	Standard
Mg	24	22616.9	1.7	**32.0867**	1.983	6.2	mg/L	32	Standard
K	39	585.0	8.2	**1.9630**	0.088	4.5	mg/L	30	Standard
Ca	43	106.7	21.1	**11.9418**	3.627	30.4	mg/L	48	Standard
Fe	54	143.6	9.6	**0.0016**	0.013	810.1	mg/L	159	Standard
Fe	57	260.0	25.0	**0.3188**	0.279	87.4	mg/L	195	Standard
> **Sc-1**	45	37107.1	4.7				mg/L	39543	Standard
Cl	35	35012.4	2.2				ug/L	27642	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	3463.7	1.0				ug/L	1647	Standard
P	31	17954.0	1.9				ug/L	15231	Standard
S	34	2235.2	3.1				ug/L	1845	Standard
Sr	88	115.0	27.2				ug/L	70	Standard
C	12	173.3	16.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	0	Standard
Dy	164	16.2	97.3				mg/L	12	Standard
Ho-1	165	16.7	62.4				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	29468.6	2.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		90.794	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.170	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606134301SD WG574205-06
Report Date/Time: Tuesday, June 28, 2016 21:55:09
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	92.166
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	89.457
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606134301SD WG574205-06
Report Date/Time: Tuesday, June 28, 2016 21:55:09
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606073201
Sample Date/Time: Tuesday, June 28, 2016 21:56:04
Number of Replicates: 3
Autosampler Position: 334
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79995.7	1.9				ug/L	81760	Standard
	Be	9	111.7	43.0	**0.1055**	0.050	47.4	ug/L	17	Standard
	Al	27	26418031.8	2.7	**175.1861**	5.381	3.1	ug/L	1410	Standard
	Sc	45	39758.8	3.3				ug/L	39543	Standard
	Ti	47	586.7	3.3	**2.3079**	0.104	4.5	ug/L	39	Standard
	V	51	2345.3	13.2	**0.1468**	0.044	29.9	ug/L	1415	Standard
	Cr	52	26092.8	3.1	**2.8709**	0.089	3.1	ug/L	6946	Standard
	Cr	53	22536.8	6.4	**26.5559**	1.479	5.6	ug/L	328	Standard
	Mn	55	128770.6	1.1	**20.6681**	0.068	0.3	ug/L	823	Standard
	Co	59	4370.6	1.2	**0.6206**	0.009	1.4	ug/L	252	Standard
	Ni	60	4217.9	1.8	**2.2421**	0.029	1.3	ug/L	192	Standard
	Cu	65	1551.7	2.2	**0.7650**	0.010	1.3	ug/L	141	Standard
	Zn	66	53557.8	1.5	**50.3807**	0.333	0.7	ug/L	491	Standard
>	**Ge**	72	491490.5	1.0				ug/L	515087	Standard
	As	75	918.5	14.0	**0.8957**	0.112	12.5	ug/L	20	Standard
	Se	82	316.6	5.4	**2.7619**	0.129	4.7	ug/L	15	Standard
	Se-1	77	1234.4	4.8	**16.4500**	0.704	4.3	ug/L	48	Standard
>	**Ga**	71	80.0	6.3				mg/L	90	Standard
	Rb	85	7720.3	1.4				ug/L	23	Standard
	Y	89	403861.7	0.4				ug/L	448510	Standard
>	**Rh**	103	80.0	50.0				ug/L	15	Standard
	Mo	98	355.0	5.6	**0.1283**	0.008	5.9	ug/L	11	Standard
	Ag	107	141.3	9.2	**0.0074**	0.002	31.2	ug/L	134	Standard
	Cd	111	202.7	10.3	**0.1036**	0.012	11.2	mg/L	7	Standard
	Cd	114	507.6	13.6	**0.1047**	0.015	13.9	ug/L	16	Standard
>	**In**	115	544497.8	0.6				ug/L	611117	Standard
	Sn	118	1023.4	24.9	**0.0925**	0.050	53.7	ug/L	518	Standard
	Sb	123	669.1	12.6	**0.1463**	0.019	13.2	ug/L	127	Standard
	Ba	135	206763.5	2.5	**112.3722**	2.187	1.9	ug/L	45	Standard
	Ce	140	411.7	6.1				ug/L	298	Standard
>	**Tb**	159	889575.8	1.3				ug/L	977942	Standard
	Ho	165	35.0	37.8				ug/L	12	Standard
	Tl	203	749.0	2.3	**0.0953**	0.003	3.6	ug/L	31	Standard
	Tl	205	685.0	13.9	**0.1029**	0.013	12.2	ug/L	20	Standard
	Pb	206	776.0	1.1	**0.0981**	0.006	5.8	ug/L	286	Standard
	Pb	207	670.3	1.0	**0.0968**	0.005	4.8	ug/L	257	Standard
	Pb	208	2692.4	0.9	**0.1003**	0.005	4.6	ug/L	956	Standard
	U	238	1446.7	3.3	**0.2244**	0.006	2.7	ug/L	17	Standard
>	**Bi**	209	494871.3	2.5				ug/L	557080	Standard

Sample ID: L1606073201
Report Date/Time: Tuesday, June 28, 2016 21:58:21
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.6945**	18.515	173.1	mg/L	0	Standard
Mg	24	122893.2	2.3	**162.6797**	1.668	1.0	mg/L	32	Standard
K	39	163.3	4.7	**0.4080**	0.035	8.6	mg/L	30	Standard
Ca	43	160.0	24.8	**18.8525**	6.186	32.8	mg/L	48	Standard
Fe	54	210.9	15.7	**0.0529**	0.031	58.0	mg/L	159	Standard
Fe	57	311.7	8.2	**0.4207**	0.052	12.3	mg/L	195	Standard
> **Sc-1**	45	39758.8	3.3				mg/L	39543	Standard
Cl	35	36983.1	0.7				ug/L	27642	Standard
Kr	83	2.3	49.5				ug/L	2	Standard
Br	81	75233.3	4.9				ug/L	1647	Standard
P	31	24466.4	2.9				ug/L	15231	Standard
S	34	2258.5	3.0				ug/L	1845	Standard
Sr	88	378.3	24.5				ug/L	70	Standard
C	12	350.0	10.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	120.9	45.9				mg/L	12	Standard
Ho-1	165	35.0	37.8				mg/L	12	Standard
Er	166	50.0	20.0				mg/L	20	Standard
I	127	539973.2	10.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		97.842	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.419	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606073201
Report Date/Time: Tuesday, June 28, 2016 21:58:21
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		89.099
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		88.833
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606073201
Report Date/Time: Tuesday, June 28, 2016 21:58:21
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606073201PS WG574571-01
Sample Date/Time: Tuesday, June 28, 2016 21:59:15
Number of Replicates: 3
Autosampler Position: 335
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83254.8	1.5				ug/L	81760	Standard
	Be	9	50562.6	1.1	**49.9976**	0.969	1.9	ug/L	17	Standard
	Al	27	27107676.6	2.1	**172.7073**	3.970	2.3	ug/L	1410	Standard
	Sc	45	41809.3	4.0				ug/L	39543	Standard
	Ti	47	632.7	1.1	**2.4753**	0.055	2.2	ug/L	39	Standard
	V	51	341856.0	1.3	**50.8320**	0.708	1.4	ug/L	1415	Standard
	Cr	52	363663.1	1.5	**53.1642**	1.402	2.6	ug/L	6946	Standard
	Cr	53	65238.5	1.6	**76.7756**	0.750	1.0	ug/L	328	Standard
	Mn	55	438903.0	0.5	**70.1258**	2.130	3.0	ug/L	823	Standard
	Co	59	331676.0	1.2	**48.9632**	1.267	2.6	ug/L	252	Standard
	Ni	60	91787.9	1.2	**50.4717**	0.832	1.6	ug/L	192	Standard
	Cu	65	90254.4	2.3	**49.1926**	0.475	1.0	ug/L	141	Standard
	Zn	66	107316.4	1.1	**100.7774**	1.488	1.5	ug/L	491	Standard
>	**Ge**	72	496483.6	2.5				ug/L	515087	Standard
	As	75	55410.9	1.5	**51.1805**	0.552	1.1	ug/L	20	Standard
	Se	82	5728.8	0.9	**51.5315**	1.616	3.1	ug/L	15	Standard
	Se-1	77	4873.5	1.7	**66.4041**	0.559	0.8	ug/L	48	Standard
>	**Ga**	71	86.7	34.8				mg/L	90	Standard
	Rb	85	8242.2	2.9				ug/L	23	Standard
	Y	89	419418.3	1.5				ug/L	448510	Standard
>	**Rh**	103	96.7	18.2				ug/L	15	Standard
	Mo	98	373.5	3.6	**0.1304**	0.006	4.5	ug/L	11	Standard
	Ag	107	276030.5	2.3	**44.4706**	1.418	3.2	ug/L	134	Standard
	Cd	111	97190.1	1.0	**49.1769**	0.174	0.4	mg/L	7	Standard
	Cd	114	234963.3	0.8	**48.9150**	0.797	1.6	ug/L	16	Standard
>	**In**	115	564991.8	0.9				ug/L	611117	Standard
	Sn	118	1183.4	5.3	**0.1145**	0.011	9.2	ug/L	518	Standard
	Sb	123	221298.1	1.9	**50.7268**	1.390	2.7	ug/L	127	Standard
	Ba	135	302634.3	0.9	**158.5372**	1.746	1.1	ug/L	45	Standard
	Ce	140	460.0	11.5				ug/L	298	Standard
>	**Tb**	159	913205.8	1.2				ug/L	977942	Standard
	Ho	165	30.0	16.7				ug/L	12	Standard
	Tl	203	392737.2	0.1	**50.5926**	0.297	0.6	ug/L	31	Standard
	Tl	205	346685.2	1.3	**51.4296**	0.576	1.1	ug/L	20	Standard
	Pb	206	257242.2	1.9	**51.6050**	1.169	2.3	ug/L	286	Standard
	Pb	207	223447.5	1.6	**49.6960**	0.910	1.8	ug/L	257	Standard
	Pb	208	904356.3	0.4	**51.9882**	0.250	0.5	ug/L	956	Standard
	U	238	346838.0	0.8	**52.5438**	0.601	1.1	ug/L	17	Standard
>	**Bi**	209	513215.0	0.6				ug/L	557080	Standard

Sample ID: L1606073201PS WG574571-01
Report Date/Time: Tuesday, June 28, 2016 22:01:32
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**9.7143**	16.817	173.1	mg/L	0	Standard
Mg	24	126816.9	1.1	**159.8003**	7.491	4.7	mg/L	32	Standard
K	39	205.0	11.2	**0.5135**	0.066	12.9	mg/L	30	Standard
Ca	43	170.0	5.9	**19.0779**	1.160	6.1	mg/L	48	Standard
Fe	54	195.8	12.4	**0.0303**	0.016	52.3	mg/L	159	Standard
Fe	57	350.0	26.8	**0.4863**	0.252	51.8	mg/L	195	Standard
> **Sc-1**	45	41809.3	4.0				mg/L	39543	Standard
Cl	35	38499.5	1.5				ug/L	27642	Standard
Kr	83	2.7	78.1				ug/L	2	Standard
Br	81	75812.6	2.8				ug/L	1647	Standard
P	31	25067.4	2.7				ug/L	15231	Standard
S	34	2465.2	4.5				ug/L	1845	Standard
Sr	88	373.3	2.8				ug/L	70	Standard
C	12	380.0	17.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	29.2	88.7				mg/L	12	Standard
Ho-1	165	30.0	16.7				mg/L	12	Standard
Er	166	16.7	69.3				mg/L	20	Standard
I	127	574284.0	7.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.828	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.388	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606073201PS WG574571-01
Report Date/Time: Tuesday, June 28, 2016 22:01:32
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	92.452
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	92.126
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606073201PS WG574571-01
Report Date/Time: Tuesday, June 28, 2016 22:01:32
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606073202SDL WG574571-02
Sample Date/Time: Tuesday, June 28, 2016 22:02:27
Number of Replicates: 3
Autosampler Position: 336
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80721.2	1.7				ug/L	81760	Standard
	Be	9	38.3	19.9	**0.0293**	0.007	25.3	ug/L	17	Standard
	Al	27	5438081.5	1.8	**35.7222**	0.611	1.7	ug/L	1410	Standard
	Sc	45	40838.3	2.8				ug/L	39543	Standard
	Ti	47	147.3	1.4	**0.4611**	0.014	3.1	ug/L	39	Standard
	V	51	1818.8	6.0	**0.0702**	0.022	30.8	ug/L	1415	Standard
	Cr	52	11457.0	0.2	**0.6827**	0.037	5.5	ug/L	6946	Standard
	Cr	53	6112.9	0.7	**7.0202**	0.190	2.7	ug/L	328	Standard
	Mn	55	27400.4	0.7	**4.3193**	0.101	2.3	ug/L	823	Standard
	Co	59	1102.0	5.1	**0.1343**	0.005	3.9	ug/L	252	Standard
	Ni	60	1029.7	4.0	**0.4727**	0.014	2.9	ug/L	192	Standard
	Cu	65	552.3	3.6	**0.2164**	0.013	6.0	ug/L	141	Standard
	Zn	66	17277.2	1.2	**15.8333**	0.273	1.7	ug/L	491	Standard
>	**Ge**	72	487411.0	2.0				ug/L	515087	Standard
	As	75	241.2	19.0	**0.2666**	0.040	15.0	ug/L	20	Standard
	Se	82	96.0	8.9	**0.7616**	0.066	8.7	ug/L	15	Standard
	Se-1	77	318.7	3.3	**3.7527**	0.181	4.8	ug/L	48	Standard
>	**Ga**	71	61.7	24.8				mg/L	90	Standard
	Rb	85	1698.4	2.7				ug/L	23	Standard
	Y	89	428587.6	1.1				ug/L	448510	Standard
>	**Rh**	103	33.3	37.7				ug/L	15	Standard
	Mo	98	99.2	5.5	**0.0185**	0.002	11.0	ug/L	11	Standard
	Ag	107	280.0	31.3	**0.0265**	0.013	49.6	ug/L	134	Standard
	Cd	111	57.2	23.8	**0.0245**	0.006	26.0	mg/L	7	Standard
	Cd	114	159.7	10.5	**0.0266**	0.003	11.9	ug/L	16	Standard
>	**In**	115	596430.5	0.6				ug/L	611117	Standard
	Sn	118	631.7	19.2	**0.0076**	0.022	282.9	ug/L	518	Standard
	Sb	123	2037.7	21.9	**0.4301**	0.100	23.2	ug/L	127	Standard
	Ba	135	42595.5	1.1	**21.1150**	0.168	0.8	ug/L	45	Standard
	Ce	140	145.0	24.9				ug/L	298	Standard
>	**Tb**	159	930009.8	1.5				ug/L	977942	Standard
	Ho	165	10.0					ug/L	12	Standard
	Tl	203	287.0	3.9	**0.0311**	0.001	3.3	ug/L	31	Standard
	Tl	205	238.3	17.8	**0.0321**	0.006	20.0	ug/L	20	Standard
	Pb	206	467.7	4.1	**0.0276**	0.003	10.6	ug/L	286	Standard
	Pb	207	403.3	2.0	**0.0292**	0.001	2.3	ug/L	257	Standard
	Pb	208	1685.7	2.5	**0.0338**	0.002	5.7	ug/L	956	Standard
	U	238	344.0	11.7	**0.0476**	0.005	11.0	ug/L	17	Standard
>	**Bi**	209	528241.9	1.3				ug/L	557080	Standard

Sample ID: L1606073202SDL WG574571-02
Report Date/Time: Tuesday, June 28, 2016 22:04:44
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	25095.8	1.4	**32.3042**	0.912	2.8	mg/L	32	Standard
K	39	53.3	28.6	**0.0336**	0.053	158.1	mg/L	30	Standard
Ca	43	51.7	14.8	**2.2359**	0.907	40.5	mg/L	48	Standard
Fe	54	140.6	10.4	**-0.0136**	0.016	114.2	mg/L	159	Standard
Fe	57	253.3	13.1	**0.2033**	0.104	51.3	mg/L	195	Standard
> **Sc-1**	45	40838.3	2.8				mg/L	39543	Standard
Cl	35	38490.2	0.9				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	17116.3	3.1				ug/L	1647	Standard
P	31	18993.6	2.2				ug/L	15231	Standard
S	34	2316.8	7.4				ug/L	1845	Standard
Sr	88	121.7	17.1				ug/L	70	Standard
C	12	150.0	13.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	5.2	126.3				mg/L	12	Standard
Ho-1	165	10.0					mg/L	12	Standard
Er	166	30.0	57.7				mg/L	20	Standard
I	127	133767.9	0.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		98.729	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.627	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606073202SDL WG574571-02
Report Date/Time: Tuesday, June 28, 2016 22:04:44
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.597
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.823
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606073202SDL WG574571-02
Sample Date/Time: Tuesday, June 28, 2016 22:05:38
Number of Replicates: 3
Autosampler Position: 337
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76053.8	2.7				ug/L	81760	Standard
	Be	9	25.0	34.6	**0.0174**	0.009	54.5	ug/L	17	Standard
	Al	27	1128796.7	0.4	**7.8615**	0.198	2.5	ug/L	1410	Standard
	Sc	45	38686.0	0.8				ug/L	39543	Standard
	Ti	47	55.3	34.5	**0.0617**	0.078	127.2	ug/L	39	Standard
	V	51	1638.9	5.5	**0.0337**	0.013	38.4	ug/L	1415	Standard
	Cr	52	8717.5	0.5	**0.2193**	0.011	4.9	ug/L	6946	Standard
	Cr	53	2070.1	6.6	**2.0621**	0.160	7.7	ug/L	328	Standard
	Mn	55	6629.8	1.0	**0.8938**	0.008	0.9	ug/L	823	Standard
	Co	59	385.0	3.7	**0.0245**	0.002	8.1	ug/L	252	Standard
	Ni	60	384.7	1.7	**0.1032**	0.003	2.7	ug/L	192	Standard
	Cu	65	339.7	2.9	**0.0913**	0.005	5.2	ug/L	141	Standard
	Zn	66	9883.2	1.3	**8.3627**	0.143	1.7	ug/L	491	Standard
>	**Ge**	72	505388.6	0.3				ug/L	515087	Standard
	As	75	-13.0	254.9	**0.0286**	0.030	104.7	ug/L	20	Standard
	Se	82	27.4	29.7	**0.1233**	0.072	58.7	ug/L	15	Standard
	Se-1	77	108.0	8.5	**0.7425**	0.129	17.3	ug/L	48	Standard
>	**Ga**	71	48.3	46.6				mg/L	90	Standard
	Rb	85	370.0	8.2				ug/L	23	Standard
	Y	89	428119.2	2.9				ug/L	448510	Standard
>	**Rh**	103	23.3	44.6				ug/L	15	Standard
	Mo	98	45.2	4.1	**-0.0015**	0.001	60.8	ug/L	11	Standard
	Ag	107	113.3	16.4	**0.0016**	0.003	172.3	ug/L	134	Standard
	Cd	111	14.3	24.6	**0.0041**	0.002	39.1	mg/L	7	Standard
	Cd	114	49.9	58.5	**0.0052**	0.006	113.8	ug/L	16	Standard
>	**In**	115	579695.7	1.7				ug/L	611117	Standard
	Sn	118	626.7	2.3	**0.0098**	0.004	44.1	ug/L	518	Standard
	Sb	123	500.2	33.4	**0.0994**	0.039	39.6	ug/L	127	Standard
	Ba	135	8786.5	0.8	**4.4621**	0.051	1.1	ug/L	45	Standard
	Ce	140	53.3	19.5				ug/L	298	Standard
>	**Tb**	159	899038.1	0.5				ug/L	977942	Standard
	Ho	165	3.3	86.6				ug/L	12	Standard
	Tl	203	124.3	3.7	**0.0110**	0.001	5.1	ug/L	31	Standard
	Tl	205	108.3	13.3	**0.0136**	0.002	15.1	ug/L	20	Standard
	Pb	206	338.3	5.2	**0.0034**	0.004	109.0	ug/L	286	Standard
	Pb	207	309.3	7.0	**0.0098**	0.005	45.9	ug/L	257	Standard
	Pb	208	1266.0	2.0	**0.0114**	0.001	11.9	ug/L	956	Standard
	U	238	68.0	9.6	**0.0072**	0.001	13.2	ug/L	17	Standard
>	**Bi**	209	520681.9	0.4				ug/L	557080	Standard

Sample ID: L1606073202SDL WG574571-02
Report Date/Time: Tuesday, June 28, 2016 22:07:55
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	5104.2	1.7	**6.8887**	0.170	2.5	mg/L	32	Standard
K	39	40.0	37.5	**-0.0032**	0.052	1630.6	mg/L	30	Standard
Ca	43	30.0	57.7	**-0.6841**	2.704	395.3	mg/L	48	Standard
Fe	54	149.1	25.4	**0.0007**	0.034	4947.2	mg/L	159	Standard
Fe	57	220.0	14.9	**0.1340**	0.108	80.4	mg/L	195	Standard
Sc-1	45	38686.0	0.8				mg/L	39543	Standard
Cl	35	39148.5	1.2				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	4957.5	4.0				ug/L	1647	Standard
P	31	18357.8	2.1				ug/L	15231	Standard
S	34	2318.5	4.4				ug/L	1845	Standard
Sr	88	78.3	18.4				ug/L	70	Standard
C	12	130.0	27.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.2	116.5				mg/L	12	Standard
Ho-1	165	3.3	86.6				mg/L	12	Standard
Er	166	16.7	91.7				mg/L	20	Standard
I	127	32523.3	5.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		93.021	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		98.117	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606073202SDL WG574571-02
Report Date/Time: Tuesday, June 28, 2016 22:07:55
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.858
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		93.466
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 22:08:51
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74027.2	2.8				ug/L	81760	Standard
	Be	9	46885.8	2.2	**52.1458**	1.090	2.1	ug/L	17	Standard
	Al	27	7258519.9	1.3	**52.0333**	2.108	4.1	ug/L	1410	Standard
	Sc	45	37711.9	1.9				ug/L	39543	Standard
	Ti	47	24276.8	0.9	**101.7561**	0.915	0.9	ug/L	39	Standard
	V	51	338117.2	0.9	**50.5811**	0.467	0.9	ug/L	1415	Standard
	Cr	52	344759.2	1.1	**50.6515**	0.612	1.2	ug/L	6946	Standard
	Cr	53	43998.8	1.8	**51.9886**	0.950	1.8	ug/L	328	Standard
	Mn	55	319756.8	1.7	**51.3453**	0.844	1.6	ug/L	823	Standard
	Co	59	342187.7	2.1	**50.8146**	1.023	2.0	ug/L	252	Standard
	Ni	60	91463.9	0.8	**50.5986**	0.401	0.8	ug/L	192	Standard
	Cu	65	92539.7	0.2	**50.7558**	0.132	0.3	ug/L	141	Standard
	Zn	66	54475.1	1.3	**51.0616**	0.732	1.4	ug/L	491	Standard
>	**Ge**	72	493368.3	0.1				ug/L	515087	Standard
	As	75	54361.0	0.4	**50.5191**	0.195	0.4	ug/L	20	Standard
	Se	82	5570.4	0.7	**50.3948**	0.327	0.6	ug/L	15	Standard
	Se-1	77	3889.2	2.7	**53.1789**	1.513	2.8	ug/L	48	Standard
>	**Ga**	71	56.7	13.5				mg/L	90	Standard
	Rb	85	710.0	9.2				ug/L	23	Standard
	Y	89	427120.1	2.2				ug/L	448510	Standard
>	**Rh**	103	20.0	50.0				ug/L	15	Standard
	Mo	98	255476.4	1.0	**100.3761**	0.479	0.5	ug/L	11	Standard
	Ag	107	311158.3	0.6	**49.2259**	0.757	1.5	ug/L	134	Standard
	Cd	111	98620.7	1.2	**49.0052**	0.618	1.3	mg/L	7	Standard
	Cd	114	238558.2	0.5	**48.7672**	0.251	0.5	ug/L	16	Standard
>	**In**	115	575337.5	0.9				ug/L	611117	Standard
	Sn	118	277257.2	0.9	**49.7064**	0.586	1.2	ug/L	518	Standard
	Sb	123	223054.4	0.8	**50.2070**	0.812	1.6	ug/L	127	Standard
	Ba	135	94954.3	0.7	**48.8337**	0.782	1.6	ug/L	45	Standard
	Ce	140	61.7	16.9				ug/L	298	Standard
>	**Tb**	159	913449.5	0.5				ug/L	977942	Standard
	Ho	165	10.0	86.6				ug/L	12	Standard
	Tl	203	397249.4	0.6	**50.3639**	0.352	0.7	ug/L	31	Standard
	Tl	205	354959.8	0.3	**51.8245**	0.065	0.1	ug/L	20	Standard
	Pb	206	256290.8	0.6	**50.5971**	0.273	0.5	ug/L	286	Standard
	Pb	207	233188.9	1.2	**51.0418**	0.565	1.1	ug/L	257	Standard
	Pb	208	935979.9	1.1	**52.9561**	0.623	1.2	ug/L	956	Standard
	U	238	344850.7	0.5	**51.4151**	0.322	0.6	ug/L	17	Standard
>	**Bi**	209	521458.9	0.2				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:11:08
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3802.1	3.0	**5.2525**	0.237	4.5	mg/L	32	Standard
K	39	1311.7	9.2	**4.5102**	0.493	10.9	mg/L	30	Standard
Ca	43	78.3	7.4	**7.1225**	0.723	10.1	mg/L	48	Standard
Fe	54	5407.8	4.5	**4.9980**	0.312	6.2	mg/L	159	Standard
Fe	57	1643.4	4.4	**5.1911**	0.156	3.0	mg/L	195	Standard
> **Sc-1**	45	37711.9	1.9				mg/L	39543	Standard
Cl	35	36184.5	2.9				ug/L	27642	Standard
Kr	83	3.0	66.7				ug/L	2	Standard
Br	81	1690.1	2.4				ug/L	1647	Standard
P	31	19537.6	2.2				ug/L	15231	Standard
S	34	2665.2	5.3				ug/L	1845	Standard
Sr	88	80.0	12.5				ug/L	70	Standard
C	12	116.7	9.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	9.4	109.4				mg/L	12	Standard
Ho-1	165	10.0	86.6				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	5821.1	15.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	104.292		
	Al	27	104.067		
	Sc	45			
	Ti	47	101.756		
	V	51	101.162		
	Cr	52	101.303		
	Cr	53			
	Mn	55	102.691		
	Co	59	101.629		
	Ni	60	101.197		
	Cu	65	101.512		
	Zn	66	102.123		
>	Ge	72		95.784	
	As	75	101.038		
	Se	82	100.790		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:11:08
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	100.376		
Ag	107	98.452		
Cd	111	98.010		
Cd	114			
In	115			94.145
Sn	118	99.413		
Sb	123	100.414		
Ba	135	97.667		
Ce	140			
Tb	159			
Ho	165			
Tl	203	100.728		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.912		
U	238	102.830		
Bi	209			93.606
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:11:08
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 22:12:02
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72720.8	1.4				ug/L	81760	Standard
	Be	9	31.7	9.1	**0.0261**	0.003	13.4	ug/L	17	Standard
	Al	27	3567.1	29.0	**0.0117**	0.008	67.2	ug/L	1410	Standard
	Sc	45	39132.1	1.1				ug/L	39543	Standard
	Ti	47	37.0	21.5	**-0.0116**	0.034	290.1	ug/L	39	Standard
	V	51	1125.9	7.5	**-0.0395**	0.013	33.1	ug/L	1415	Standard
	Cr	52	6280.0	4.7	**-0.1269**	0.043	33.9	ug/L	6946	Standard
	Cr	53	665.0	3.8	**0.4397**	0.027	6.1	ug/L	328	Standard
	Mn	55	716.0	7.9	**-0.0346**	0.009	26.9	ug/L	823	Standard
	Co	59	261.7	5.4	**0.0070**	0.002	26.9	ug/L	252	Standard
	Ni	60	220.7	25.5	**0.0158**	0.031	194.2	ug/L	192	Standard
	Cu	65	175.7	31.3	**0.0045**	0.030	664.0	ug/L	141	Standard
	Zn	66	494.7	12.2	**-0.3623**	0.056	15.6	ug/L	491	Standard
>	**Ge**	72	499522.9	0.5				ug/L	515087	Standard
	As	75	-22.4	199.4	**0.0197**	0.041	208.7	ug/L	20	Standard
	Se	82	18.1	44.1	**0.0430**	0.071	165.8	ug/L	15	Standard
	Se-1	77	56.7	13.4	**0.0566**	0.102	180.9	ug/L	48	Standard
>	**Ga**	71	45.0	38.5				mg/L	90	Standard
	Rb	85	31.7	45.6				ug/L	23	Standard
	Y	89	417360.5	0.9				ug/L	448510	Standard
>	**Rh**	103	6.7	114.6				ug/L	15	Standard
	Mo	98	287.6	23.9	**0.0927**	0.026	27.9	ug/L	11	Standard
	Ag	107	190.3	4.8	**0.0136**	0.002	11.9	ug/L	134	Standard
	Cd	111	25.8	62.0	**0.0098**	0.008	81.4	mg/L	7	Standard
	Cd	114	75.5	64.7	**0.0104**	0.010	96.5	ug/L	16	Standard
>	**In**	115	581386.0	1.0				ug/L	611117	Standard
	Sn	118	788.4	29.5	**0.0382**	0.041	108.2	ug/L	518	Standard
	Sb	123	553.2	30.8	**0.1105**	0.038	34.6	ug/L	127	Standard
	Ba	135	104.3	62.1	**0.0282**	0.034	119.2	ug/L	45	Standard
	Ce	140	15.0	88.2				ug/L	298	Standard
>	**Tb**	159	916886.1	2.0				ug/L	977942	Standard
	Ho	165	6.7	43.3				ug/L	12	Standard
	Tl	203	96.0	40.0	**0.0072**	0.005	66.3	ug/L	31	Standard
	Tl	205	80.0	25.0	**0.0092**	0.003	29.7	ug/L	20	Standard
	Pb	206	283.3	0.7	**-0.0085**	0.001	14.2	ug/L	286	Standard
	Pb	207	242.3	7.5	**-0.0058**	0.004	64.9	ug/L	257	Standard
	Pb	208	1066.3	4.7	**-0.0011**	0.002	176.0	ug/L	956	Standard
	U	238	68.3	24.9	**0.0070**	0.002	33.1	ug/L	17	Standard
>	**Bi**	209	530213.7	1.5				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:14:19
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**32.2313**	32.195	99.9	mg/L	0	Standard
Mg	24	43.3	46.6	**0.0012**	0.027	2331.0	mg/L	32	Standard
K	39	38.3	41.9	**-0.0103**	0.056	538.4	mg/L	30	Standard
Ca	43	25.0	20.0	**-1.5137**	0.732	48.3	mg/L	48	Standard
Fe	54	136.0	18.8	**-0.0125**	0.025	197.8	mg/L	159	Standard
Fe	57	235.0	17.0	**0.1764**	0.130	73.9	mg/L	195	Standard
Sc-1	45	39132.1	1.1				mg/L	39543	Standard
Cl	35	35003.7	1.3				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	1530.1	14.2				ug/L	1647	Standard
P	31	18486.3	1.6				ug/L	15231	Standard
S	34	2496.9	3.2				ug/L	1845	Standard
Sr	88	51.7	31.1				ug/L	70	Standard
C	12	86.7	52.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.5	5.0				mg/L	12	Standard
Ho-1	165	6.7	43.3				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	4170.6	14.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.978	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:14:19
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.135
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.177
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:14:19
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606073202
Sample Date/Time: Tuesday, June 28, 2016 22:15:14
Number of Replicates: 3
Autosampler Position: 338
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76318.4	2.0				ug/L	81760	Standard
	Be	9	40.0	69.6	**0.0330**	0.029	88.9	ug/L	17	Standard
	Al	27	23571551.0	1.4	**163.8296**	2.147	1.3	ug/L	1410	Standard
	Sc	45	39020.2	3.0				ug/L	39543	Standard
	Ti	47	595.0	5.6	**2.3771**	0.145	6.1	ug/L	39	Standard
	V	51	1578.4	25.4	**0.0346**	0.061	176.4	ug/L	1415	Standard
	Cr	52	21713.5	1.1	**2.2571**	0.032	1.4	ug/L	6946	Standard
	Cr	53	25227.7	4.6	**30.1930**	1.400	4.6	ug/L	328	Standard
	Mn	55	113835.4	1.3	**18.5078**	0.247	1.3	ug/L	823	Standard
	Co	59	3991.2	2.2	**0.5722**	0.015	2.6	ug/L	252	Standard
	Ni	60	3905.8	1.2	**2.0985**	0.020	1.0	ug/L	192	Standard
	Cu	65	1542.7	2.9	**0.7718**	0.023	3.0	ug/L	141	Standard
	Zn	66	25635.3	1.1	**24.0224**	0.322	1.3	ug/L	491	Standard
>	**Ge**	72	484802.3	0.3				ug/L	515087	Standard
	As	75	911.4	10.9	**0.9016**	0.094	10.4	ug/L	20	Standard
	Se	82	285.6	1.4	**2.5161**	0.032	1.3	ug/L	15	Standard
	Se-1	77	1182.4	4.2	**15.9563**	0.701	4.4	ug/L	48	Standard
>	**Ga**	71	88.3	32.7				mg/L	90	Standard
	Rb	85	8777.5	4.2				ug/L	23	Standard
	Y	89	401864.9	0.9				ug/L	448510	Standard
>	**Rh**	103	95.0	19.0				ug/L	15	Standard
	Mo	98	789.3	4.8	**0.3176**	0.018	5.8	ug/L	11	Standard
	Ag	107	101.3	14.3	**0.0012**	0.002	197.6	ug/L	134	Standard
	Cd	111	81.6	8.8	**0.0411**	0.003	8.4	mg/L	7	Standard
	Cd	114	236.0	13.0	**0.0475**	0.007	15.0	ug/L	16	Standard
>	**In**	115	530144.7	1.0				ug/L	611117	Standard
	Sn	118	770.0	10.1	**0.0481**	0.014	28.6	ug/L	518	Standard
	Sb	123	728.9	13.4	**0.1651**	0.022	13.6	ug/L	127	Standard
	Ba	135	185424.0	0.3	**103.5140**	0.966	0.9	ug/L	45	Standard
	Ce	140	496.7	18.3				ug/L	298	Standard
>	**Tb**	159	867434.9	1.9				ug/L	977942	Standard
	Ho	165	20.0	0.0				ug/L	12	Standard
	Tl	203	239.3	4.0	**0.0278**	0.001	4.3	ug/L	31	Standard
	Tl	205	225.0	20.4	**0.0330**	0.007	22.5	ug/L	20	Standard
	Pb	206	582.0	1.5	**0.0598**	0.001	1.4	ug/L	286	Standard
	Pb	207	480.3	3.9	**0.0549**	0.005	9.4	ug/L	257	Standard
	Pb	208	1935.7	1.6	**0.0572**	0.003	4.8	ug/L	956	Standard
	U	238	1057.0	6.3	**0.1661**	0.010	5.7	ug/L	17	Standard
>	**Bi**	209	486031.1	0.8				ug/L	557080	Standard

Sample ID: L1606073202
Report Date/Time: Tuesday, June 28, 2016 22:17:31
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	123145.1	1.2	**166.1283**	3.309	2.0	mg/L	32	Standard
K	39	193.3	27.9	**0.5225**	0.193	37.0	mg/L	30	Standard
Ca	43	155.0	19.6	**18.5842**	5.220	28.1	mg/L	48	Standard
Fe	54	201.4	15.9	**0.0482**	0.034	70.8	mg/L	159	Standard
Fe	57	301.7	7.8	**0.4095**	0.108	26.4	mg/L	195	Standard
Sc-1	45	39020.2	3.0				mg/L	39543	Standard
Cl	35	36882.8	1.2				ug/L	27642	Standard
Kr	83	2.3	65.5				ug/L	2	Standard
Br	81	66321.6	2.9				ug/L	1647	Standard
P	31	23950.6	0.8				ug/L	15231	Standard
S	34	2285.2	3.2				ug/L	1845	Standard
Sr	88	326.7	6.4				ug/L	70	Standard
C	12	383.3	11.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	24.1	87.6				mg/L	12	Standard
Ho-1	165	20.0	0.0				mg/L	12	Standard
Er	166	53.3	28.6				mg/L	20	Standard
I	127	490270.4	8.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		93.344	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		94.121	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606073202
Report Date/Time: Tuesday, June 28, 2016 22:17:31
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	86.750
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	87.246
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606073202
Report Date/Time: Tuesday, June 28, 2016 22:17:31
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606076201
Sample Date/Time: Tuesday, June 28, 2016 22:18:25
Number of Replicates: 3
Autosampler Position: 339
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74725.5	1.6				ug/L	81760	Standard
	Be	9	33.3	56.8	**0.0272**	0.022	79.2	ug/L	17	Standard
	Al	27	1272215.4	1.6	**9.0180**	0.230	2.5	ug/L	1410	Standard
	Sc	45	40079.6	0.9				ug/L	39543	Standard
	Ti	47	73.0	21.4	**0.1312**	0.065	49.6	ug/L	39	Standard
	V	51	1912.1	11.7	**0.0708**	0.033	46.3	ug/L	1415	Standard
	Cr	52	9243.8	1.7	**0.2820**	0.015	5.4	ug/L	6946	Standard
	Cr	53	1550.1	16.6	**1.4381**	0.284	19.7	ug/L	328	Standard
	Mn	55	4050.6	51.7	**0.4818**	0.327	67.8	ug/L	823	Standard
	Co	59	576.0	42.1	**0.0513**	0.035	68.0	ug/L	252	Standard
	Ni	60	659.3	6.6	**0.2481**	0.023	9.5	ug/L	192	Standard
	Cu	65	493.7	5.0	**0.1712**	0.013	7.5	ug/L	141	Standard
	Zn	66	2294.8	5.3	**1.2839**	0.114	8.9	ug/L	491	Standard
>	**Ge**	72	510757.4	0.7				ug/L	515087	Standard
	As	75	-15.4	127.0	**0.0265**	0.018	66.6	ug/L	20	Standard
	Se	82	23.9	38.0	**0.0895**	0.079	87.8	ug/L	15	Standard
	Se-1	77	93.7	16.1	**0.5342**	0.193	36.2	ug/L	48	Standard
>	**Ga**	71	58.3	17.8				mg/L	90	Standard
	Rb	85	1476.7	6.6				ug/L	23	Standard
	Y	89	432514.2	1.7				ug/L	448510	Standard
>	**Rh**	103	26.7	39.0				ug/L	15	Standard
	Mo	98	146.0	48.2	**0.0371**	0.027	72.5	ug/L	11	Standard
	Ag	107	112.7	13.4	**0.0013**	0.002	181.1	ug/L	134	Standard
	Cd	111	43.2	14.9	**0.0182**	0.003	18.1	mg/L	7	Standard
	Cd	114	109.3	34.7	**0.0170**	0.007	43.8	ug/L	16	Standard
>	**In**	115	586396.8	0.9				ug/L	611117	Standard
	Sn	118	771.7	15.6	**0.0341**	0.022	63.3	ug/L	518	Standard
	Sb	123	236.4	15.7	**0.0394**	0.008	21.0	ug/L	127	Standard
	Ba	135	3173.0	2.9	**1.5769**	0.056	3.5	ug/L	45	Standard
	Ce	140	2270.2	3.3				ug/L	298	Standard
>	**Tb**	159	916697.6	0.7				ug/L	977942	Standard
	Ho	165	490.0	12.4				ug/L	12	Standard
	Tl	203	130.7	10.3	**0.0114**	0.002	13.6	ug/L	31	Standard
	Tl	205	215.0	77.4	**0.0282**	0.023	82.9	ug/L	20	Standard
	Pb	206	395.3	3.6	**0.0126**	0.003	22.5	ug/L	286	Standard
	Pb	207	368.7	19.2	**0.0207**	0.015	70.3	ug/L	257	Standard
	Pb	208	1630.7	27.1	**0.0295**	0.024	80.4	ug/L	956	Standard
	U	238	158.7	69.5	**0.0200**	0.016	79.0	ug/L	17	Standard
>	**Bi**	209	534609.1	0.7				ug/L	557080	Standard

Sample ID: L1606076201
Report Date/Time: Tuesday, June 28, 2016 22:20:42
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**31.5470**	31.355	99.4	mg/L	0	Standard
Mg	24	598.3	11.5	**0.7287**	0.091	12.5	mg/L	32	Standard
K	39	176.7	18.2	**0.4475**	0.107	24.0	mg/L	30	Standard
Ca	43	93.3	48.6	**8.6233**	6.782	78.6	mg/L	48	Standard
Fe	54	114.2	15.8	**-0.0351**	0.016	46.3	mg/L	159	Standard
Fe	57	271.7	7.4	**0.2806**	0.074	26.3	mg/L	195	Standard
> **Sc-1**	45	40079.6	0.9				mg/L	39543	Standard
Cl	35	37059.3	1.9				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	3193.7	9.3				ug/L	1647	Standard
P	31	19133.7	1.0				ug/L	15231	Standard
S	34	2435.2	3.0				ug/L	1845	Standard
Sr	88	116.7	19.8				ug/L	70	Standard
C	12	110.0	24.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	699.7	7.7				mg/L	12	Standard
Ho-1	165	490.0	12.4				mg/L	12	Standard
Er	166	563.3	4.5				mg/L	20	Standard
I	127	24296.3	37.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.396	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.159	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606076201
Report Date/Time: Tuesday, June 28, 2016 22:20:42
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.955
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.966
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606138001
Sample Date/Time: Tuesday, June 28, 2016 22:21:37
Number of Replicates: 3
Autosampler Position: 340
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79317.1	2.1				ug/L	81760	Standard
	Be	9	923.4	3.6	**0.9491**	0.046	4.8	ug/L	17	Standard
	Al	27	22322355.7	1.2	**149.2842**	1.816	1.2	ug/L	1410	Standard
	Sc	45	50147.9	2.6				ug/L	39543	Standard
	Ti	47	26223.7	0.7	**107.9697**	0.846	0.8	ug/L	39	Standard
	V	51	127306.3	1.5	**18.5751**	0.291	1.6	ug/L	1415	Standard
	Cr	52	100654.9	1.1	**13.7705**	0.145	1.1	ug/L	6946	Standard
	Cr	53	12865.5	4.5	**14.6884**	0.707	4.8	ug/L	328	Standard
	Mn	55	4304305.9	0.4	**680.6871**	4.381	0.6	ug/L	823	Standard
	Co	59	49359.3	0.1	**7.1725**	0.024	0.3	ug/L	252	Standard
	Ni	60	30538.1	1.5	**16.5228**	0.280	1.7	ug/L	192	Standard
	Cu	65	24616.6	0.6	**13.1942**	0.101	0.8	ug/L	141	Standard
	Zn	66	126791.0	1.1	**117.7932**	1.321	1.1	ug/L	491	Standard
>	**Ge**	72	502314.8	0.3				ug/L	515087	Standard
	As	75	6358.7	2.6	**5.8397**	0.152	2.6	ug/L	20	Standard
	Se	82	52.0	15.4	**0.3436**	0.072	20.9	ug/L	15	Standard
	Se-1	77	111.0	6.5	**0.7919**	0.097	12.2	ug/L	48	Standard
>	**Ga**	71	15155.9	0.2				mg/L	90	Standard
	Rb	85	244761.5	1.6				ug/L	23	Standard
	Y	89	581808.2	2.1				ug/L	448510	Standard
>	**Rh**	103	81.7	28.3				ug/L	15	Standard
	Mo	98	7400.1	1.4	**2.8571**	0.099	3.5	ug/L	11	Standard
	Ag	107	359.7	8.1	**0.0401**	0.005	12.7	ug/L	134	Standard
	Cd	111	226.0	4.7	**0.1082**	0.006	6.0	mg/L	7	Standard
	Cd	114	572.5	10.7	**0.1110**	0.014	12.9	ug/L	16	Standard
>	**In**	115	581977.2	2.1				ug/L	611117	Standard
	Sn	118	1150.0	8.4	**0.1021**	0.013	12.6	ug/L	518	Standard
	Sb	123	1041.4	3.0	**0.2191**	0.011	5.1	ug/L	127	Standard
	Ba	135	264934.9	0.4	**134.7601**	2.267	1.7	ug/L	45	Standard
	Ce	140	872470.7	0.6				ug/L	298	Standard
>	**Tb**	159	944063.2	1.7				ug/L	977942	Standard
	Ho	165	13182.4	3.3				ug/L	12	Standard
	Tl	203	1500.7	3.5	**0.1808**	0.008	4.2	ug/L	31	Standard
	Tl	205	1268.4	3.4	**0.1783**	0.007	4.2	ug/L	20	Standard
	Pb	206	85066.5	1.3	**16.3322**	0.328	2.0	ug/L	286	Standard
	Pb	207	70090.8	1.2	**14.9184**	0.287	1.9	ug/L	257	Standard
	Pb	208	293052.1	0.7	**16.1237**	0.113	0.7	ug/L	956	Standard
	U	238	6791.5	1.5	**0.9844**	0.021	2.1	ug/L	17	Standard
>	**Bi**	209	534845.3	0.8				ug/L	557080	Standard

Sample ID: L1606138001
Report Date/Time: Tuesday, June 28, 2016 22:23:54
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	2491.9	2.3	**2.5590**	0.057	2.2	mg/L	32	Standard
K	39	860.0	5.6	**2.1528**	0.183	8.5	mg/L	30	Standard
Ca	43	481.7	3.2	**52.3892**	3.048	5.8	mg/L	48	Standard
Fe	54	14060.9	4.3	**9.8944**	0.202	2.0	mg/L	159	Standard
Fe	57	4057.2	4.0	**10.1835**	0.599	5.9	mg/L	195	Standard
> **Sc-1**	45	50147.9	2.6				mg/L	39543	Standard
Cl	35	34061.6	1.3				ug/L	27642	Standard
Kr	83	2.7	57.3				ug/L	2	Standard
Br	81	3443.7	4.1				ug/L	1647	Standard
P	31	33871.2	1.5				ug/L	15231	Standard
S	34	2461.9	2.1				ug/L	1845	Standard
Sr	88	305.0	8.2				ug/L	70	Standard
C	12	206.7	29.6				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	40.0	50.0				mg/L	0	Standard
Dy	164	22121.1	1.8				mg/L	12	Standard
Ho-1	165	13182.4	3.3				mg/L	12	Standard
Er	166	12021.4	0.9				mg/L	20	Standard
I	127	158658.4	11.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		97.012	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.520	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138001
Report Date/Time: Tuesday, June 28, 2016 22:23:54
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	95.232
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	96.009
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Ti 47 Upper, S, EEE	Ti	47	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606138001
Report Date/Time: Tuesday, June 28, 2016 22:23:54
Page 3

Approved: June 29, 2016

Zn 66 Upper, S, EEE	Zn	66
Ba 135 Upper, S, EEE	Ba	135

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606138002
Sample Date/Time: Tuesday, June 28, 2016 22:24:48
Number of Replicates: 3
Autosampler Position: 341
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80508.4	1.6				ug/L	81760	Standard
	Be	9	865.0	8.5	**0.8749**	0.075	8.6	ug/L	17	Standard
	Al	27	22868958.8	1.7	**150.6704**	2.885	1.9	ug/L	1410	Standard
	Sc	45	49400.4	2.8				ug/L	39543	Standard
	Ti	47	23164.7	1.4	**96.5938**	1.528	1.6	ug/L	39	Standard
	V	51	124883.8	1.8	**18.4559**	0.304	1.6	ug/L	1415	Standard
	Cr	52	103466.8	1.3	**14.3825**	0.239	1.7	ug/L	6946	Standard
	Cr	53	12758.7	1.5	**14.7559**	0.244	1.7	ug/L	328	Standard
	Mn	55	4407919.2	0.3	**706.0993**	3.755	0.5	ug/L	823	Standard
	Co	59	51228.1	1.3	**7.5421**	0.111	1.5	ug/L	252	Standard
	Ni	60	30969.0	1.2	**16.9756**	0.235	1.4	ug/L	192	Standard
	Cu	65	25034.7	0.7	**13.5947**	0.118	0.9	ug/L	141	Standard
	Zn	66	152662.0	1.2	**143.8470**	1.906	1.3	ug/L	491	Standard
>	**Ge**	72	495895.8	0.2				ug/L	515087	Standard
	As	75	6401.6	1.0	**5.9544**	0.051	0.9	ug/L	20	Standard
	Se	82	65.6	3.6	**0.4728**	0.020	4.3	ug/L	15	Standard
	Se-1	77	120.3	12.9	**0.9403**	0.215	22.9	ug/L	48	Standard
>	**Ga**	71	15282.7	1.5				mg/L	90	Standard
	Rb	85	248664.6	2.3				ug/L	23	Standard
	Y	89	598252.6	1.6				ug/L	448510	Standard
>	**Rh**	103	61.7	4.7				ug/L	15	Standard
	Mo	98	8173.4	1.7	**3.0442**	0.032	1.1	ug/L	11	Standard
	Ag	107	334.0	5.8	**0.0342**	0.003	8.9	ug/L	134	Standard
	Cd	111	175.2	4.5	**0.0802**	0.004	5.0	mg/L	7	Standard
	Cd	114	424.7	8.5	**0.0779**	0.007	8.4	ug/L	16	Standard
>	**In**	115	603234.5	0.7				ug/L	611117	Standard
	Sn	118	1006.7	2.2	**0.0706**	0.004	6.0	ug/L	518	Standard
	Sb	123	981.2	4.8	**0.1979**	0.011	5.8	ug/L	127	Standard
	Ba	135	265403.3	0.1	**130.2141**	0.915	0.7	ug/L	45	Standard
	Ce	140	904460.6	1.2				ug/L	298	Standard
>	**Tb**	159	953156.2	1.3				ug/L	977942	Standard
	Ho	165	13287.5	2.7				ug/L	12	Standard
	Tl	203	1338.1	1.2	**0.1623**	0.004	2.2	ug/L	31	Standard
	Tl	205	1380.1	21.0	**0.1966**	0.045	23.0	ug/L	20	Standard
	Pb	206	85143.9	1.2	**16.5139**	0.424	2.6	ug/L	286	Standard
	Pb	207	70337.2	0.8	**15.1235**	0.335	2.2	ug/L	257	Standard
	Pb	208	299122.5	0.5	**16.6275**	0.345	2.1	ug/L	956	Standard
	U	238	7037.6	3.1	**1.0308**	0.048	4.7	ug/L	17	Standard
>	**Bi**	209	529543.6	1.6				ug/L	557080	Standard

Sample ID: L1606138002
Report Date/Time: Tuesday, June 28, 2016 22:27:05
Page 1

Approved: June 29, 2016

Na	23	5.0	0.0	**25.7128**	0.720	2.8	mg/L	0	Standard
Mg	24	2751.9	1.6	**2.8760**	0.054	1.9	mg/L	32	Standard
K	39	820.0	12.4	**2.0772**	0.274	13.2	mg/L	30	Standard
Ca	43	460.0	15.1	**50.7618**	10.089	19.9	mg/L	48	Standard
Fe	54	14800.2	1.5	**10.5924**	0.445	4.2	mg/L	159	Standard
Fe	57	4213.9	2.7	**10.7740**	0.623	5.8	mg/L	195	Standard
Sc-1	45	49400.4	2.8				mg/L	39543	Standard
Cl	35	34900.8	0.4				ug/L	27642	Standard
Kr	83	3.0	0.0				ug/L	2	Standard
Br	81	3303.7	3.6				ug/L	1647	Standard
P	31	33101.1	2.4				ug/L	15231	Standard
S	34	2478.5	3.3				ug/L	1845	Standard
Sr	88	288.3	13.9				ug/L	70	Standard
C	12	206.7	17.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	53.3	65.8				mg/L	0	Standard
Dy	164	23538.0	3.9				mg/L	12	Standard
Ho-1	165	13287.5	2.7				mg/L	12	Standard
Er	166	12271.6	4.5				mg/L	20	Standard
I	127	174817.7	8.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		98.469	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.274	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138002
Report Date/Time: Tuesday, June 28, 2016 22:27:05
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.710
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.057
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606138002
Report Date/Time: Tuesday, June 28, 2016 22:27:05
Page 3

Approved: June 29, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606138106
Sample Date/Time: Tuesday, June 28, 2016 22:28:00
Number of Replicates: 3
Autosampler Position: 342
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75884.6	2.0				ug/L	81760	Standard
	Be	9	26.7	10.8	**0.0192**	0.003	17.5	ug/L	17	Standard
	Al	27	26057.5	7.4	**0.1676**	0.010	6.0	ug/L	1410	Standard
	Sc	45	39473.0	0.8				ug/L	39543	Standard
	Ti	47	39.3	21.6	**-0.0051**	0.035	681.9	ug/L	39	Standard
	V	51	1622.1	7.1	**0.0295**	0.017	56.8	ug/L	1415	Standard
	Cr	52	8616.4	0.2	**0.1952**	0.003	1.5	ug/L	6946	Standard
	Cr	53	775.0	5.7	**0.5518**	0.051	9.2	ug/L	328	Standard
	Mn	55	5073.2	5.2	**0.6432**	0.042	6.5	ug/L	823	Standard
	Co	59	305.3	2.9	**0.0126**	0.001	10.1	ug/L	252	Standard
	Ni	60	336.7	8.9	**0.0759**	0.016	21.2	ug/L	192	Standard
	Cu	65	1953.8	0.6	**0.9495**	0.006	0.6	ug/L	141	Standard
	Zn	66	2327.2	0.6	**1.3204**	0.013	1.0	ug/L	491	Standard
>	**Ge**	72	509122.8	0.0				ug/L	515087	Standard
	As	75	-29.7	15.9	**0.0136**	0.004	31.1	ug/L	20	Standard
	Se	82	19.1	15.2	**0.0489**	0.026	52.3	ug/L	15	Standard
	Se-1	77	54.3	12.9	**0.0108**	0.094	874.4	ug/L	48	Standard
>	**Ga**	71	61.7	23.4				mg/L	90	Standard
	Rb	85	193.3	9.8				ug/L	23	Standard
	Y	89	438080.9	2.7				ug/L	448510	Standard
>	**Rh**	103	13.3	94.4				ug/L	15	Standard
	Mo	98	1348.1	2.9	**0.4917**	0.018	3.6	ug/L	11	Standard
	Ag	107	96.0	6.3	**-0.0016**	0.001	59.2	ug/L	134	Standard
	Cd	111	29.0	10.6	**0.0110**	0.001	13.2	mg/L	7	Standard
	Cd	114	62.4	58.3	**0.0074**	0.007	97.2	ug/L	16	Standard
>	**In**	115	596825.6	0.5				ug/L	611117	Standard
	Sn	118	1153.4	11.3	**0.0978**	0.023	23.4	ug/L	518	Standard
	Sb	123	161.5	18.7	**0.0222**	0.007	29.7	ug/L	127	Standard
	Ba	135	505.0	6.7	**0.2253**	0.016	7.0	ug/L	45	Standard
	Ce	140	233.3	21.5				ug/L	298	Standard
>	**Tb**	159	918727.2	0.8				ug/L	977942	Standard
	Ho	165	6.7	43.3				ug/L	12	Standard
	Tl	203	77.0	15.3	**0.0048**	0.001	30.3	ug/L	31	Standard
	Tl	205	70.0	7.1	**0.0078**	0.001	10.7	ug/L	20	Standard
	Pb	206	353.0	8.2	**0.0052**	0.006	124.2	ug/L	286	Standard
	Pb	207	316.0	5.0	**0.0102**	0.004	41.5	ug/L	257	Standard
	Pb	208	1279.0	2.9	**0.0109**	0.002	20.1	ug/L	956	Standard
	U	238	66.0	9.1	**0.0067**	0.001	13.4	ug/L	17	Standard
>	**Bi**	209	529518.6	1.3				ug/L	557080	Standard

Sample ID: L1606138106
Report Date/Time: Tuesday, June 28, 2016 22:30:17
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.6931**	18.512	173.1	mg/L	0	Standard
Mg	24	105.0	31.2	**0.0830**	0.044	52.9	mg/L	32	Standard
K	39	31.7	9.1	**-0.0342**	0.011	31.1	mg/L	30	Standard
Ca	43	30.0	66.7	**-0.7662**	3.081	402.2	mg/L	48	Standard
Fe	54	132.0	13.9	**-0.0174**	0.016	93.3	mg/L	159	Standard
Fe	57	243.3	16.5	**0.1990**	0.142	71.4	mg/L	195	Standard
Sc-1	45	39473.0	0.8				mg/L	39543	Standard
Cl	35	35269.0	1.4				ug/L	27642	Standard
Kr	83	2.3	107.9				ug/L	2	Standard
Br	81	2306.8	8.4				ug/L	1647	Standard
P	31	16594.1	0.5				ug/L	15231	Standard
S	34	2131.8	6.4				ug/L	1845	Standard
Sr	88	88.3	36.8				ug/L	70	Standard
C	12	163.3	23.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	9.0	110.6				mg/L	12	Standard
Ho-1	165	6.7	43.3				mg/L	12	Standard
Er	166	20.0	0.0				mg/L	20	Standard
I	127	8462.7	47.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		92.814	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		98.842	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138106
Report Date/Time: Tuesday, June 28, 2016 22:30:17
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.661
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.052
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139301
Sample Date/Time: Tuesday, June 28, 2016 22:31:11
Number of Replicates: 3
Autosampler Position: 343
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75653.4	0.9				ug/L	81760	Standard
	Be	9	36.7	15.7	**0.0301**	0.006	19.8	ug/L	17	Standard
	Al	27	730309.3	1.1	**5.1065**	0.100	1.9	ug/L	1410	Standard
	Sc	45	40455.6	1.8				ug/L	39543	Standard
	Ti	47	3049.6	0.8	**12.0214**	0.208	1.7	ug/L	39	Standard
	V	51	20044.5	0.4	**2.6590**	0.030	1.1	ug/L	1415	Standard
	Cr	52	14783.9	1.2	**1.0533**	0.031	2.9	ug/L	6946	Standard
	Cr	53	1555.1	3.4	**1.4185**	0.043	3.0	ug/L	328	Standard
	Mn	55	57374.1	1.8	**8.6446**	0.070	0.8	ug/L	823	Standard
	Co	59	1706.1	0.5	**0.2099**	0.002	1.1	ug/L	252	Standard
	Ni	60	1958.1	2.2	**0.9282**	0.028	3.1	ug/L	192	Standard
	Cu	65	11994.7	1.9	**6.1812**	0.063	1.0	ug/L	141	Standard
	Zn	66	20161.1	0.6	**17.4500**	0.168	1.0	ug/L	491	Standard
>	**Ge**	72	518413.1	1.4				ug/L	515087	Standard
	As	75	157.8	25.1	**0.1801**	0.037	20.6	ug/L	20	Standard
	Se	82	23.1	48.8	**0.0801**	0.098	121.9	ug/L	15	Standard
	Se-1	77	63.3	7.3	**0.1169**	0.069	59.2	ug/L	48	Standard
>	**Ga**	71	505.0	4.5				mg/L	90	Standard
	Rb	85	7842.0	2.0				ug/L	23	Standard
	Y	89	431479.5	0.9				ug/L	448510	Standard
>	**Rh**	103	15.0	100.0				ug/L	15	Standard
	Mo	98	468.3	2.4	**0.1592**	0.005	3.4	ug/L	11	Standard
	Ag	107	177.3	7.3	**0.0110**	0.002	18.6	ug/L	134	Standard
	Cd	111	203.8	5.3	**0.0952**	0.005	5.7	mg/L	7	Standard
	Cd	114	524.4	13.4	**0.0989**	0.013	13.5	ug/L	16	Standard
>	**In**	115	593880.8	0.6				ug/L	611117	Standard
	Sn	118	2320.2	4.6	**0.3018**	0.016	5.3	ug/L	518	Standard
	Sb	123	1520.6	3.7	**0.3188**	0.011	3.4	ug/L	127	Standard
	Ba	135	21231.5	1.4	**10.5571**	0.086	0.8	ug/L	45	Standard
	Ce	140	21453.5	1.1				ug/L	298	Standard
>	**Tb**	159	941666.6	1.6				ug/L	977942	Standard
	Ho	165	511.7	11.3				ug/L	12	Standard
	Tl	203	106.0	9.9	**0.0082**	0.001	16.0	ug/L	31	Standard
	Tl	205	85.0	31.1	**0.0097**	0.004	38.6	ug/L	20	Standard
	Pb	206	17358.9	1.4	**3.2466**	0.044	1.3	ug/L	286	Standard
	Pb	207	14395.2	1.0	**2.9852**	0.035	1.2	ug/L	257	Standard
	Pb	208	59327.6	0.5	**3.1814**	0.019	0.6	ug/L	956	Standard
	U	238	205.3	8.1	**0.0266**	0.002	9.0	ug/L	17	Standard
>	**Bi**	209	540531.5	0.2				ug/L	557080	Standard

Sample ID: L1606139301
Report Date/Time: Tuesday, June 28, 2016 22:33:28
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**21.0062**	36.375	173.2	mg/L	0	Standard
Mg	24	250.0	13.1	**0.2677**	0.037	13.9	mg/L	32	Standard
K	39	110.0	28.4	**0.2220**	0.104	47.1	mg/L	30	Standard
Ca	43	51.7	20.1	**2.3448**	1.674	71.4	mg/L	48	Standard
Fe	54	595.3	12.6	**0.3888**	0.057	14.6	mg/L	159	Standard
Fe	57	388.3	11.5	**0.6585**	0.168	25.5	mg/L	195	Standard
Sc-1	45	40455.6	1.8				mg/L	39543	Standard
Cl	35	34669.0	0.4				ug/L	27642	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	3693.8	7.2				ug/L	1647	Standard
P	31	17358.3	1.4				ug/L	15231	Standard
S	34	2180.2	7.0				ug/L	1845	Standard
Sr	88	80.0	38.0				ug/L	70	Standard
C	12	186.7	21.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	0	Standard
Dy	164	723.1	8.6				mg/L	12	Standard
Ho-1	165	511.7	11.3				mg/L	12	Standard
Er	166	493.3	12.2				mg/L	20	Standard
I	127	46309.3	8.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		92.531	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		100.646	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606139301
Report Date/Time: Tuesday, June 28, 2016 22:33:28
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.180
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.029
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139302
Sample Date/Time: Tuesday, June 28, 2016 22:34:23
Number of Replicates: 3
Autosampler Position: 344
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76487.6	0.6				ug/L	81760	Standard
	Be	9	26.7	57.3	**0.0190**	0.016	86.6	ug/L	17	Standard
	Al	27	422387.9	1.8	**2.9148**	0.060	2.0	ug/L	1410	Standard
	Sc	45	40846.7	1.8				ug/L	39543	Standard
	Ti	47	1809.8	2.0	**7.1474**	0.208	2.9	ug/L	39	Standard
	V	51	13250.7	0.8	**1.7088**	0.032	1.8	ug/L	1415	Standard
	Cr	52	14384.8	1.8	**1.0188**	0.008	0.8	ug/L	6946	Standard
	Cr	53	1510.1	5.3	**1.3859**	0.061	4.4	ug/L	328	Standard
	Mn	55	44577.2	1.1	**6.7600**	0.108	1.6	ug/L	823	Standard
	Co	59	1155.0	3.6	**0.1339**	0.008	5.8	ug/L	252	Standard
	Ni	60	2034.1	3.4	**0.9799**	0.027	2.8	ug/L	192	Standard
	Cu	65	13331.2	1.6	**6.9601**	0.242	3.5	ug/L	141	Standard
	Zn	66	18863.7	1.1	**16.4635**	0.229	1.4	ug/L	491	Standard
>	**Ge**	72	512762.7	2.0				ug/L	515087	Standard
	As	75	120.3	20.8	**0.1481**	0.025	16.6	ug/L	20	Standard
	Se	82	22.7	50.1	**0.0800**	0.103	128.5	ug/L	15	Standard
	Se-1	77	66.7	3.8	**0.1706**	0.046	27.2	ug/L	48	Standard
>	**Ga**	71	313.3	7.5				mg/L	90	Standard
	Rb	85	9062.7	6.6				ug/L	23	Standard
	Y	89	438095.5	1.1				ug/L	448510	Standard
>	**Rh**	103	11.7	65.5				ug/L	15	Standard
	Mo	98	749.7	5.2	**0.2622**	0.015	5.9	ug/L	11	Standard
	Ag	107	250.0	4.9	**0.0215**	0.002	9.3	ug/L	134	Standard
	Cd	111	268.7	5.4	**0.1246**	0.007	5.7	mg/L	7	Standard
	Cd	114	686.3	8.9	**0.1290**	0.012	9.4	ug/L	16	Standard
>	**In**	115	602735.5	0.4				ug/L	611117	Standard
	Sn	118	2001.8	9.6	**0.2412**	0.032	13.1	ug/L	518	Standard
	Sb	123	1551.7	1.3	**0.3206**	0.005	1.5	ug/L	127	Standard
	Ba	135	10171.4	0.7	**4.9702**	0.026	0.5	ug/L	45	Standard
	Ce	140	10203.4	3.4				ug/L	298	Standard
>	**Tb**	159	943711.5	1.3				ug/L	977942	Standard
	Ho	165	201.7	2.9				ug/L	12	Standard
	Tl	203	136.7	14.9	**0.0117**	0.003	21.7	ug/L	31	Standard
	Tl	205	126.7	4.6	**0.0153**	0.001	5.0	ug/L	20	Standard
	Pb	206	13666.8	0.5	**2.5119**	0.014	0.6	ug/L	286	Standard
	Pb	207	11351.3	1.7	**2.3133**	0.023	1.0	ug/L	257	Standard
	Pb	208	47777.6	0.4	**2.5197**	0.028	1.1	ug/L	956	Standard
	U	238	179.7	5.5	**0.0226**	0.001	5.2	ug/L	17	Standard
>	**Bi**	209	546981.3	0.8				ug/L	557080	Standard

Sample ID: L1606139302
Report Date/Time: Tuesday, June 28, 2016 22:36:40
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	296.7	10.2	**0.3250**	0.036	11.0	mg/L	32	Standard
K	39	131.7	25.3	**0.2905**	0.116	40.0	mg/L	30	Standard
Ca	43	28.3	44.4	**-1.1569**	1.919	165.9	mg/L	48	Standard
Fe	54	412.0	10.8	**0.2238**	0.038	16.8	mg/L	159	Standard
Fe	57	300.0	5.8	**0.3560**	0.058	16.2	mg/L	195	Standard
> **Sc-1**	45	40846.7	1.8				mg/L	39543	Standard
Cl	35	35351.2	2.0				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	2080.1	6.1				ug/L	1647	Standard
P	31	17501.8	0.9				ug/L	15231	Standard
S	34	2256.8	4.2				ug/L	1845	Standard
Sr	88	83.3	24.2				ug/L	70	Standard
C	12	216.7	41.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	30.0	66.7				mg/L	0	Standard
Dy	164	374.4	8.4				mg/L	12	Standard
Ho-1	165	201.7	2.9				mg/L	12	Standard
Er	166	186.7	17.2				mg/L	20	Standard
I	127	9588.0	8.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.551	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.549	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606139302
Report Date/Time: Tuesday, June 28, 2016 22:36:40
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.629
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.187
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139303
Sample Date/Time: Tuesday, June 28, 2016 22:37:34
Number of Replicates: 3
Autosampler Position: 345
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75224.7	2.8				ug/L	81760	Standard
	Be	9	26.7	47.2	**0.0195**	0.014	72.7	ug/L	17	Standard
	Al	27	335427.7	1.4	**2.3514**	0.039	1.7	ug/L	1410	Standard
	Sc	45	40873.4	3.2				ug/L	39543	Standard
	Ti	47	1306.7	3.3	**5.2014**	0.227	4.4	ug/L	39	Standard
	V	51	16739.3	1.3	**2.2527**	0.061	2.7	ug/L	1415	Standard
	Cr	52	12212.3	0.6	**0.7342**	0.019	2.6	ug/L	6946	Standard
	Cr	53	1208.4	9.2	**1.0637**	0.123	11.6	ug/L	328	Standard
	Mn	55	30356.4	1.8	**4.6331**	0.125	2.7	ug/L	823	Standard
	Co	59	971.0	0.7	**0.1098**	0.003	2.3	ug/L	252	Standard
	Ni	60	1423.4	2.9	**0.6666**	0.013	2.0	ug/L	192	Standard
	Cu	65	8494.4	2.1	**4.4744**	0.140	3.1	ug/L	141	Standard
	Zn	66	12099.5	0.6	**10.4446**	0.200	1.9	ug/L	491	Standard
>	**Ge**	72	504506.4	1.3				ug/L	515087	Standard
	As	75	95.4	18.1	**0.1269**	0.016	12.6	ug/L	20	Standard
	Se	82	22.1	16.3	**0.0773**	0.034	44.4	ug/L	15	Standard
	Se-1	77	59.3	9.3	**0.0847**	0.065	76.5	ug/L	48	Standard
>	**Ga**	71	286.7	13.3				mg/L	90	Standard
	Rb	85	4860.8	3.1				ug/L	23	Standard
	Y	89	447487.9	1.5				ug/L	448510	Standard
>	**Rh**	103	10.0	50.0				ug/L	15	Standard
	Mo	98	335.7	2.2	**0.1056**	0.003	3.1	ug/L	11	Standard
	Ag	107	149.3	7.8	**0.0061**	0.002	27.6	ug/L	134	Standard
	Cd	111	92.8	5.9	**0.0407**	0.003	6.7	mg/L	7	Standard
	Cd	114	275.4	18.0	**0.0483**	0.009	19.4	ug/L	16	Standard
>	**In**	115	608945.1	0.4				ug/L	611117	Standard
	Sn	118	2160.2	6.9	**0.2646**	0.024	9.0	ug/L	518	Standard
	Sb	123	1475.9	3.1	**0.3011**	0.011	3.5	ug/L	127	Standard
	Ba	135	7298.8	1.0	**3.5230**	0.049	1.4	ug/L	45	Standard
	Ce	140	7186.7	3.0				ug/L	298	Standard
>	**Tb**	159	941207.0	0.6				ug/L	977942	Standard
	Ho	165	175.0	29.8				ug/L	12	Standard
	Tl	203	127.3	10.7	**0.0108**	0.002	16.0	ug/L	31	Standard
	Tl	205	91.7	22.0	**0.0106**	0.003	26.7	ug/L	20	Standard
	Pb	206	10382.9	0.6	**1.9175**	0.012	0.6	ug/L	286	Standard
	Pb	207	8724.8	0.9	**1.7876**	0.028	1.6	ug/L	257	Standard
	Pb	208	36599.7	1.1	**1.9407**	0.035	1.8	ug/L	956	Standard
	U	238	138.7	9.6	**0.0170**	0.002	11.7	ug/L	17	Standard
>	**Bi**	209	540229.2	0.6				ug/L	557080	Standard

Sample ID: L1606139303
Report Date/Time: Tuesday, June 28, 2016 22:39:51
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	201.7	23.0	**0.2027**	0.061	30.0	mg/L	32	Standard
K	39	85.0	15.6	**0.1365**	0.043	31.8	mg/L	30	Standard
Ca	43	46.7	34.4	**1.4659**	2.107	143.8	mg/L	48	Standard
Fe	54	348.6	7.5	**0.1680**	0.018	11.0	mg/L	159	Standard
Fe	57	325.0	9.4	**0.4364**	0.086	19.7	mg/L	195	Standard
> **Sc-1**	45	40873.4	3.2				mg/L	39543	Standard
Cl	35	35045.8	0.8				ug/L	27642	Standard
Kr	83	2.3	107.9				ug/L	2	Standard
Br	81	1873.4	9.4				ug/L	1647	Standard
P	31	17319.9	1.5				ug/L	15231	Standard
S	34	2126.8	6.5				ug/L	1845	Standard
Sr	88	78.3	28.8				ug/L	70	Standard
C	12	160.0	25.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	78.1				mg/L	0	Standard
Dy	164	223.9	25.3				mg/L	12	Standard
Ho-1	165	175.0	29.8				mg/L	12	Standard
Er	166	126.7	25.4				mg/L	20	Standard
I	127	7363.5	5.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		92.007	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		97.946	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606139303
Report Date/Time: Tuesday, June 28, 2016 22:39:51
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.645
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.975
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139304
Sample Date/Time: Tuesday, June 28, 2016 22:40:46
Number of Replicates: 3
Autosampler Position: 346
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74628.4	2.8				ug/L	81760	Standard
	Be	9	20.0	25.0	**0.0122**	0.005	40.5	ug/L	17	Standard
	Al	27	327323.4	5.6	**2.3109**	0.073	3.1	ug/L	1410	Standard
	Sc	45	39232.4	1.5				ug/L	39543	Standard
	Ti	47	1081.7	2.8	**4.2827**	0.197	4.6	ug/L	39	Standard
	V	51	17364.8	3.6	**2.3462**	0.053	2.3	ug/L	1415	Standard
	Cr	52	11969.4	3.2	**0.6998**	0.018	2.6	ug/L	6946	Standard
	Cr	53	1151.7	1.8	**0.9999**	0.045	4.5	ug/L	328	Standard
	Mn	55	27710.7	5.0	**4.2196**	0.194	4.6	ug/L	823	Standard
	Co	59	805.7	4.4	**0.0858**	0.004	4.6	ug/L	252	Standard
	Ni	60	1156.4	2.7	**0.5226**	0.014	2.7	ug/L	192	Standard
	Cu	65	7318.8	3.6	**3.8448**	0.072	1.9	ug/L	141	Standard
	Zn	66	9717.1	4.5	**8.2301**	0.259	3.2	ug/L	491	Standard
>	**Ge**	72	504100.5	3.0				ug/L	515087	Standard
	As	75	86.2	19.2	**0.1185**	0.013	11.2	ug/L	20	Standard
	Se	82	21.0	26.2	**0.0675**	0.052	77.2	ug/L	15	Standard
	Se-1	77	57.7	9.6	**0.0648**	0.094	144.5	ug/L	48	Standard
>	**Ga**	71	233.3	13.9				mg/L	90	Standard
	Rb	85	3882.2	6.3				ug/L	23	Standard
	Y	89	430223.1	5.6				ug/L	448510	Standard
>	**Rh**	103	8.3	124.9				ug/L	15	Standard
	Mo	98	801.5	6.3	**0.2860**	0.012	4.3	ug/L	11	Standard
	Ag	107	120.7	5.8	**0.0023**	0.000	18.3	ug/L	134	Standard
	Cd	111	92.6	9.2	**0.0417**	0.004	9.3	mg/L	7	Standard
	Cd	114	241.3	24.7	**0.0431**	0.013	29.2	ug/L	16	Standard
>	**In**	115	593856.9	3.7				ug/L	611117	Standard
	Sn	118	1755.1	4.0	**0.2036**	0.011	5.3	ug/L	518	Standard
	Sb	123	1169.8	3.7	**0.2423**	0.004	1.7	ug/L	127	Standard
	Ba	135	5956.2	4.1	**2.9440**	0.065	2.2	ug/L	45	Standard
	Ce	140	5319.3	7.3				ug/L	298	Standard
>	**Tb**	159	910819.3	2.0				ug/L	977942	Standard
	Ho	165	128.3	12.5				ug/L	12	Standard
	Tl	203	88.0	5.7	**0.0062**	0.001	12.2	ug/L	31	Standard
	Tl	205	70.0	25.8	**0.0077**	0.003	33.5	ug/L	20	Standard
	Pb	206	8718.5	3.6	**1.6271**	0.072	4.4	ug/L	286	Standard
	Pb	207	7163.0	4.3	**1.4818**	0.072	4.9	ug/L	257	Standard
	Pb	208	29853.9	4.1	**1.5981**	0.065	4.1	ug/L	956	Standard
	U	238	115.3	2.6	**0.0139**	0.000	1.6	ug/L	17	Standard
>	**Bi**	209	531672.0	1.7				ug/L	557080	Standard

Sample ID: L1606139304
Report Date/Time: Tuesday, June 28, 2016 22:43:03
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	220.0	2.3	**0.2381**	0.006	2.4	mg/L	32	Standard
K	39	73.3	3.9	**0.1083**	0.007	6.4	mg/L	30	Standard
Ca	43	31.7	24.1	**-0.5049**	1.115	220.9	mg/L	48	Standard
Fe	54	263.9	12.4	**0.1035**	0.028	27.1	mg/L	159	Standard
Fe	57	248.3	16.1	**0.2213**	0.145	65.6	mg/L	195	Standard
> **Sc-1**	45	39232.4	1.5				mg/L	39543	Standard
Cl	35	34033.6	5.9				ug/L	27642	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	1946.8	12.8				ug/L	1647	Standard
P	31	16857.7	2.7				ug/L	15231	Standard
S	34	2100.1	4.5				ug/L	1845	Standard
Sr	88	66.7	28.4				ug/L	70	Standard
C	12	176.7	23.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0					mg/L	0	Standard
Dy	164	134.9	39.4				mg/L	12	Standard
Ho-1	165	128.3	12.5				mg/L	12	Standard
Er	166	106.7	10.8				mg/L	20	Standard
I	127	6791.5	2.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.277	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.867	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606139304
Report Date/Time: Tuesday, June 28, 2016 22:43:03
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.176
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.439
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139305
Sample Date/Time: Tuesday, June 28, 2016 22:43:57
Number of Replicates: 3
Autosampler Position: 347
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74899.7	0.9				ug/L	81760	Standard
	Be	9	18.3	15.7	**0.0104**	0.003	29.0	ug/L	17	Standard
	Al	27	328437.1	1.5	**2.3115**	0.035	1.5	ug/L	1410	Standard
	Sc	45	40647.8	2.1				ug/L	39543	Standard
	Ti	47	1117.7	1.6	**4.2600**	0.050	1.2	ug/L	39	Standard
	V	51	13332.6	0.8	**1.6821**	0.023	1.4	ug/L	1415	Standard
	Cr	52	11640.1	0.5	**0.5893**	0.015	2.6	ug/L	6946	Standard
	Cr	53	1158.4	10.6	**0.9575**	0.132	13.8	ug/L	328	Standard
	Mn	55	25508.1	0.7	**3.7255**	0.023	0.6	ug/L	823	Standard
	Co	59	795.0	2.3	**0.0800**	0.002	2.0	ug/L	252	Standard
	Ni	60	1111.4	4.5	**0.4759**	0.022	4.7	ug/L	192	Standard
	Cu	65	7781.0	2.0	**3.9419**	0.127	3.2	ug/L	141	Standard
	Zn	66	11686.8	2.5	**9.6690**	0.148	1.5	ug/L	491	Standard
>	**Ge**	72	523177.1	1.2				ug/L	515087	Standard
	As	75	150.7	20.5	**0.1724**	0.029	16.5	ug/L	20	Standard
	Se	82	42.9	7.7	**0.2478**	0.029	11.9	ug/L	15	Standard
	Se-1	77	60.7	2.5	**0.0739**	0.016	21.3	ug/L	48	Standard
>	**Ga**	71	226.7	1.3				mg/L	90	Standard
	Rb	85	5315.9	7.9				ug/L	23	Standard
	Y	89	431953.8	1.9				ug/L	448510	Standard
>	**Rh**	103	15.0	66.7				ug/L	15	Standard
	Mo	98	276.3	6.3	**0.0857**	0.006	6.9	ug/L	11	Standard
	Ag	107	169.0	7.2	**0.0096**	0.002	19.7	ug/L	134	Standard
	Cd	111	162.4	6.5	**0.0750**	0.005	6.0	mg/L	7	Standard
	Cd	114	335.5	8.9	**0.0613**	0.005	8.9	ug/L	16	Standard
>	**In**	115	595959.3	0.7				ug/L	611117	Standard
	Sn	118	1471.7	5.0	**0.1534**	0.014	9.4	ug/L	518	Standard
	Sb	123	1173.4	5.5	**0.2422**	0.014	5.7	ug/L	127	Standard
	Ba	135	5837.1	3.3	**2.8740**	0.084	2.9	ug/L	45	Standard
	Ce	140	5496.0	2.0				ug/L	298	Standard
>	**Tb**	159	937149.5	1.0				ug/L	977942	Standard
	Ho	165	138.3	7.5				ug/L	12	Standard
	Tl	203	94.3	16.1	**0.0066**	0.002	26.5	ug/L	31	Standard
	Tl	205	66.7	26.3	**0.0070**	0.002	32.8	ug/L	20	Standard
	Pb	206	8082.5	0.7	**1.4592**	0.036	2.4	ug/L	286	Standard
	Pb	207	6688.5	1.5	**1.3390**	0.041	3.1	ug/L	257	Standard
	Pb	208	27392.7	1.4	**1.4181**	0.004	0.2	ug/L	956	Standard
	U	238	108.7	8.4	**0.0125**	0.002	12.3	ug/L	17	Standard
>	**Bi**	209	547233.0	1.6				ug/L	557080	Standard

Sample ID: L1606139305
Report Date/Time: Tuesday, June 28, 2016 22:46:14
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.3347**	17.892	173.1	mg/L	0	Standard
Mg	24	221.7	15.4	**0.2295**	0.039	17.1	mg/L	32	Standard
K	39	100.0	17.3	**0.1871**	0.055	29.3	mg/L	30	Standard
Ca	43	26.7	60.3	**-1.4304**	2.300	160.8	mg/L	48	Standard
Fe	54	289.4	9.6	**0.1178**	0.025	21.2	mg/L	159	Standard
Fe	57	271.7	9.3	**0.2686**	0.097	36.1	mg/L	195	Standard
> **Sc-1**	45	40647.8	2.1				mg/L	39543	Standard
Cl	35	33690.1	1.5				ug/L	27642	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	7261.7	1.9				ug/L	1647	Standard
P	31	16325.5	0.4				ug/L	15231	Standard
S	34	1950.1	10.4				ug/L	1845	Standard
Sr	88	81.7	9.4				ug/L	70	Standard
C	12	186.7	18.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	222.3	16.3				mg/L	12	Standard
Ho-1	165	138.3	7.5				mg/L	12	Standard
Er	166	160.0	6.3				mg/L	20	Standard
I	127	6493.1	4.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		91.609	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		101.571	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606139305
Report Date/Time: Tuesday, June 28, 2016 22:46:14
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.520
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.232
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 22:47:10
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	71793.2	3.7				ug/L	81760	Standard
	Be	9	44618.9	1.2	**51.1958**	1.696	3.3	ug/L	17	Standard
	Al	27	6865175.8	2.0	**50.7482**	1.963	3.9	ug/L	1410	Standard
	Sc	45	39468.0	3.1				ug/L	39543	Standard
	Ti	47	23906.5	1.3	**96.7870**	2.402	2.5	ug/L	39	Standard
	V	51	334011.6	0.6	**48.2557**	0.941	2.0	ug/L	1415	Standard
	Cr	52	340377.2	0.5	**48.2561**	1.001	2.1	ug/L	6946	Standard
	Cr	53	42902.3	1.0	**48.9484**	1.164	2.4	ug/L	328	Standard
	Mn	55	309303.1	1.2	**47.9606**	0.591	1.2	ug/L	823	Standard
	Co	59	332037.9	0.8	**47.6215**	0.384	0.8	ug/L	252	Standard
	Ni	60	90431.4	0.3	**48.3175**	0.794	1.6	ug/L	192	Standard
	Cu	65	91394.5	0.3	**48.4155**	0.851	1.8	ug/L	141	Standard
	Zn	66	53482.9	1.3	**48.3851**	1.341	2.8	ug/L	491	Standard
>	Ge	72	510859.4	1.5				ug/L	515087	Standard
	As	75	54007.5	0.9	**48.4846**	1.105	2.3	ug/L	20	Standard
	Se	82	5360.1	0.8	**46.8280**	0.601	1.3	ug/L	15	Standard
	Se-1	77	3686.4	1.7	**48.6230**	0.812	1.7	ug/L	48	Standard
>	Ga	71	56.7	18.4				mg/L	90	Standard
	Rb	85	810.0	4.5				ug/L	23	Standard
	Y	89	431931.2	2.6				ug/L	448510	Standard
>	Rh	103	26.7	43.3				ug/L	15	Standard
	Mo	98	251225.1	0.7	**96.8468**	1.378	1.4	ug/L	11	Standard
	Ag	107	304440.8	0.4	**47.2505**	0.474	1.0	ug/L	134	Standard
	Cd	111	96060.5	0.9	**46.8341**	0.834	1.8	mg/L	7	Standard
	Cd	114	238889.8	1.2	**47.9103**	0.301	0.6	ug/L	16	Standard
>	In	115	586435.3	1.3				ug/L	611117	Standard
	Sn	118	269259.0	2.4	**47.3660**	1.700	3.6	ug/L	518	Standard
	Sb	123	219432.6	0.9	**48.4541**	0.360	0.7	ug/L	127	Standard
	Ba	135	92563.9	0.7	**46.7027**	0.660	1.4	ug/L	45	Standard
	Ce	140	88.3	61.1				ug/L	298	Standard
>	Tb	159	918919.9	0.8				ug/L	977942	Standard
	Ho	165	13.3	21.7				ug/L	12	Standard
	Tl	203	386832.4	0.9	**48.5637**	0.777	1.6	ug/L	31	Standard
	Tl	205	343277.1	1.0	**49.6247**	0.132	0.3	ug/L	20	Standard
	Pb	206	249448.7	0.4	**48.7606**	0.291	0.6	ug/L	286	Standard
	Pb	207	228146.2	0.6	**49.4471**	0.531	1.1	ug/L	257	Standard
	Pb	208	902842.6	0.2	**50.5774**	0.397	0.8	ug/L	956	Standard
	U	238	333363.3	0.9	**49.2162**	0.753	1.5	ug/L	17	Standard
>	Bi	209	526646.5	0.9				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:49:27
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.4874**	18.156	173.1	mg/L	0	Standard
Mg	24	3567.1	3.5	**4.7067**	0.317	6.7	mg/L	32	Standard
K	39	1421.7	5.7	**4.6773**	0.373	8.0	mg/L	30	Standard
Ca	43	60.0	8.3	**3.7779**	0.573	15.2	mg/L	48	Standard
Fe	54	5303.1	4.9	**4.6695**	0.087	1.9	mg/L	159	Standard
Fe	57	1570.1	3.9	**4.6916**	0.328	7.0	mg/L	195	Standard
Sc-1	45	39468.0	3.1				mg/L	39543	Standard
Cl	35	34178.5	1.9				ug/L	27642	Standard
Kr	83	1.7	124.9				ug/L	2	Standard
Br	81	1696.8	1.5				ug/L	1647	Standard
P	31	17313.2	2.3				ug/L	15231	Standard
S	34	2535.2	5.3				ug/L	1845	Standard
Sr	88	81.7	40.8				ug/L	70	Standard
C	12	120.0	14.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	8.9	11.2				mg/L	12	Standard
Ho-1	165	13.3	21.7				mg/L	12	Standard
Er	166	23.3	89.2				mg/L	20	Standard
I	127	2468.5	18.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	102.392		
Al	27	101.496		
Sc	45			
Ti	47	96.787		
V	51	96.511		
Cr	52	96.512		
Cr	53			
Mn	55	95.921		
Co	59	95.243		
Ni	60	96.635		
Cu	65	96.831		
Zn	66	96.770		
Ge	72		99.179	
As	75	96.969		
Se	82	93.656		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:49:27
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	96.847		
Ag	107	94.501		
Cd	111	93.668		
Cd	114			
In	115			95.961
Sn	118	94.732		
Sb	123	96.908		
Ba	135	93.405		
Ce	140			
Tb	159			
Ho	165			
Tl	203	97.127		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.155		
U	238	98.432		
Bi	209			94.537
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:49:27
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 22:50:22
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72519.8	1.9				ug/L	81760	Standard
	Be	9	21.7	113.8	**0.0146**	0.027	188.1	ug/L	17	Standard
	Al	27	2256.8	25.0	**0.0021**	0.004	187.1	ug/L	1410	Standard
	Sc	45	39016.8	1.3				ug/L	39543	Standard
	Ti	47	28.7	19.2	**-0.0470**	0.020	43.5	ug/L	39	Standard
	V	51	1194.9	2.1	**-0.0301**	0.006	21.3	ug/L	1415	Standard
	Cr	52	6354.0	0.6	**-0.1199**	0.015	12.7	ug/L	6946	Standard
	Cr	53	495.0	4.6	**0.2373**	0.020	8.3	ug/L	328	Standard
	Mn	55	724.7	7.6	**-0.0338**	0.007	19.6	ug/L	823	Standard
	Co	59	253.7	9.1	**0.0057**	0.003	47.4	ug/L	252	Standard
	Ni	60	174.7	5.0	**-0.0098**	0.003	33.3	ug/L	192	Standard
	Cu	65	124.0	10.1	**-0.0239**	0.006	23.4	ug/L	141	Standard
	Zn	66	416.7	1.4	**-0.4373**	0.011	2.5	ug/L	491	Standard
>	**Ge**	72	501758.9	1.8				ug/L	515087	Standard
	As	75	-30.1	127.3	**0.0123**	0.035	285.9	ug/L	20	Standard
	Se	82	17.1	40.3	**0.0328**	0.058	178.2	ug/L	15	Standard
	Se-1	77	51.3	18.1	**-0.0207**	0.113	545.7	ug/L	48	Standard
>	**Ga**	71	38.3	27.2				mg/L	90	Standard
	Rb	85	28.3	36.7				ug/L	23	Standard
	Y	89	432168.4	0.3				ug/L	448510	Standard
>	**Rh**	103	15.0	66.7				ug/L	15	Standard
	Mo	98	274.3	30.4	**0.0862**	0.033	37.9	ug/L	11	Standard
	Ag	107	126.3	11.5	**0.0033**	0.002	71.2	ug/L	134	Standard
	Cd	111	15.5	45.6	**0.0046**	0.003	75.0	mg/L	7	Standard
	Cd	114	32.8	69.1	**0.0016**	0.005	280.8	ug/L	16	Standard
>	**In**	115	589750.3	1.0				ug/L	611117	Standard
	Sn	118	705.0	18.1	**0.0215**	0.021	98.3	ug/L	518	Standard
	Sb	123	430.7	24.1	**0.0816**	0.022	27.0	ug/L	127	Standard
	Ba	135	35.7	21.2	**-0.0072**	0.004	55.2	ug/L	45	Standard
	Ce	140	38.3	41.9				ug/L	298	Standard
>	**Tb**	159	920829.0	0.8				ug/L	977942	Standard
	Ho	165	5.0	100.0				ug/L	12	Standard
	Tl	203	59.7	35.8	**0.0027**	0.003	98.6	ug/L	31	Standard
	Tl	205	48.3	29.9	**0.0047**	0.002	43.8	ug/L	20	Standard
	Pb	206	270.0	6.9	**-0.0110**	0.004	33.8	ug/L	286	Standard
	Pb	207	257.3	3.4	**-0.0024**	0.002	65.9	ug/L	257	Standard
	Pb	208	1000.7	8.0	**-0.0046**	0.004	92.2	ug/L	956	Standard
	U	238	61.0	23.2	**0.0060**	0.002	34.0	ug/L	17	Standard
>	**Bi**	209	528739.4	0.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:52:39
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	23.3	86.6	**-0.0259**	0.027	103.6	mg/L	32	Standard
K	39	31.7	48.2	**-0.0333**	0.052	155.8	mg/L	30	Standard
Ca	43	30.0	72.6	**-0.7014**	3.391	483.5	mg/L	48	Standard
Fe	54	129.4	20.1	**-0.0182**	0.025	139.3	mg/L	159	Standard
Fe	57	250.0	16.4	**0.2316**	0.148	64.1	mg/L	195	Standard
Sc-1	45	39016.8	1.3				mg/L	39543	Standard
Cl	35	33625.3	1.5				ug/L	27642	Standard
Kr	83	3.3	124.9				ug/L	2	Standard
Br	81	1580.1	13.5				ug/L	1647	Standard
P	31	17735.4	1.8				ug/L	15231	Standard
S	34	2481.9	5.6				ug/L	1845	Standard
Sr	88	73.3	14.2				ug/L	70	Standard
C	12	113.3	41.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	9.4	5.9				mg/L	12	Standard
Ho-1	165	5.0	100.0				mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	1860.1	15.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		97.412	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:52:39
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.504
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.913
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:52:39
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139306
Sample Date/Time: Tuesday, June 28, 2016 22:53:35
Number of Replicates: 3
Autosampler Position: 348
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70220.8	3.4				ug/L	81760	Standard
	Be	9	21.7	48.0	**0.0156**	0.012	75.7	ug/L	17	Standard
	Al	27	339019.6	1.3	**2.5478**	0.066	2.6	ug/L	1410	Standard
	Sc	45	39077.0	1.3				ug/L	39543	Standard
	Ti	47	1359.1	3.4	**5.3458**	0.211	3.9	ug/L	39	Standard
	V	51	13893.8	2.6	**1.8093**	0.072	4.0	ug/L	1415	Standard
	Cr	52	10697.1	2.4	**0.4928**	0.058	11.9	ug/L	6946	Standard
	Cr	53	1106.7	6.2	**0.9306**	0.098	10.5	ug/L	328	Standard
	Mn	55	27827.9	2.7	**4.1809**	0.179	4.3	ug/L	823	Standard
	Co	59	887.0	0.6	**0.0960**	0.003	2.8	ug/L	252	Standard
	Ni	60	1041.0	6.1	**0.4528**	0.043	9.5	ug/L	192	Standard
	Cu	65	6718.2	2.4	**3.4752**	0.137	3.9	ug/L	141	Standard
	Zn	66	10492.3	2.9	**8.8264**	0.376	4.3	ug/L	491	Standard
>	**Ge**	72	510902.9	1.6				ug/L	515087	Standard
	As	75	95.4	28.2	**0.1256**	0.023	18.4	ug/L	20	Standard
	Se	82	17.2	18.1	**0.0307**	0.025	82.4	ug/L	15	Standard
	Se-1	77	52.3	18.1	**-0.0171**	0.139	811.5	ug/L	48	Standard
>	**Ga**	71	256.7	6.0				mg/L	90	Standard
	Rb	85	4037.2	4.0				ug/L	23	Standard
	Y	89	441838.7	1.2				ug/L	448510	Standard
>	**Rh**	103	13.3	43.3				ug/L	15	Standard
	Mo	98	321.3	2.1	**0.1057**	0.004	4.2	ug/L	11	Standard
	Ag	107	134.7	13.7	**0.0048**	0.003	55.1	ug/L	134	Standard
	Cd	111	94.4	13.5	**0.0434**	0.006	14.4	mg/L	7	Standard
	Cd	114	184.5	30.3	**0.0322**	0.011	33.4	ug/L	16	Standard
>	**In**	115	582517.6	1.6				ug/L	611117	Standard
	Sn	118	1633.4	8.5	**0.1881**	0.027	14.5	ug/L	518	Standard
	Sb	123	1202.8	2.0	**0.2547**	0.009	3.5	ug/L	127	Standard
	Ba	135	5625.0	0.7	**2.8338**	0.047	1.7	ug/L	45	Standard
	Ce	140	7943.7	1.7				ug/L	298	Standard
>	**Tb**	159	895525.5	0.4				ug/L	977942	Standard
	Ho	165	181.7	15.7				ug/L	12	Standard
	Tl	203	104.3	12.5	**0.0084**	0.002	20.6	ug/L	31	Standard
	Tl	205	91.7	22.0	**0.0111**	0.003	27.4	ug/L	20	Standard
	Pb	206	6727.8	0.5	**1.2603**	0.018	1.4	ug/L	286	Standard
	Pb	207	5614.0	0.6	**1.1666**	0.017	1.4	ug/L	257	Standard
	Pb	208	23664.6	2.1	**1.2740**	0.038	3.0	ug/L	956	Standard
	U	238	141.7	17.4	**0.0181**	0.004	20.7	ug/L	17	Standard
>	**Bi**	209	523874.2	0.9				ug/L	557080	Standard

Sample ID: L1606139306
Report Date/Time: Tuesday, June 28, 2016 22:55:52
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	196.7	25.5	**0.2078**	0.067	32.2	mg/L	32	Standard
K	39	58.3	32.5	**0.0586**	0.068	115.6	mg/L	30	Standard
Ca	43	33.3	48.2	**-0.2427**	2.425	998.9	mg/L	48	Standard
Fe	54	342.2	5.9	**0.1765**	0.023	12.8	mg/L	159	Standard
Fe	57	293.3	7.1	**0.3778**	0.076	20.2	mg/L	195	Standard
Sc-1	45	39077.0	1.3				mg/L	39543	Standard
Cl	35	33443.5	2.1				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	1690.1	11.7				ug/L	1647	Standard
P	31	17176.4	4.2				ug/L	15231	Standard
S	34	2306.8	4.6				ug/L	1845	Standard
Sr	88	88.3	31.2				ug/L	70	Standard
C	12	143.3	10.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	86.6				mg/L	0	Standard
Dy	164	280.6	4.7				mg/L	12	Standard
Ho-1	165	181.7	15.7				mg/L	12	Standard
Er	166	196.7	36.8				mg/L	20	Standard
I	127	4592.4	1.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		85.886	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.188	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606139306
Report Date/Time: Tuesday, June 28, 2016 22:55:52
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.320
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.039
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139307
Sample Date/Time: Tuesday, June 28, 2016 22:56:46
Number of Replicates: 3
Autosampler Position: 349
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72494.7	0.8				ug/L	81760	Standard
	Be	9	18.3	15.7	**0.0111**	0.003	30.5	ug/L	17	Standard
	Al	27	342418.0	1.1	**2.4911**	0.036	1.4	ug/L	1410	Standard
	Sc	45	40101.3	1.0				ug/L	39543	Standard
	Ti	47	932.7	2.7	**3.5634**	0.149	4.2	ug/L	39	Standard
	V	51	14612.8	2.1	**1.8826**	0.024	1.3	ug/L	1415	Standard
	Cr	52	11386.6	1.0	**0.5692**	0.057	10.1	ug/L	6946	Standard
	Cr	53	1260.1	6.9	**1.0869**	0.132	12.1	ug/L	328	Standard
	Mn	55	23640.1	1.0	**3.4768**	0.122	3.5	ug/L	823	Standard
	Co	59	741.4	0.2	**0.0735**	0.003	3.5	ug/L	252	Standard
	Ni	60	1110.0	3.0	**0.4811**	0.030	6.3	ug/L	192	Standard
	Cu	65	7534.9	0.8	**3.8515**	0.130	3.4	ug/L	141	Standard
	Zn	66	11297.5	1.4	**9.4161**	0.135	1.4	ug/L	491	Standard
>	**Ge**	72	518404.2	2.6				ug/L	515087	Standard
	As	75	185.7	13.2	**0.2041**	0.018	8.6	ug/L	20	Standard
	Se	82	58.4	11.2	**0.3845**	0.051	13.1	ug/L	15	Standard
	Se-1	77	57.0	12.3	**0.0333**	0.095	284.1	ug/L	48	Standard
>	**Ga**	71	201.7	7.6				mg/L	90	Standard
	Rb	85	5255.9	2.3				ug/L	23	Standard
	Y	89	429044.9	0.4				ug/L	448510	Standard
>	**Rh**	103	18.3	68.6				ug/L	15	Standard
	Mo	98	299.7	1.8	**0.0961**	0.002	1.9	ug/L	11	Standard
	Ag	107	153.3	22.0	**0.0075**	0.005	67.1	ug/L	134	Standard
	Cd	111	143.0	17.2	**0.0665**	0.011	16.8	mg/L	7	Standard
	Cd	114	356.8	31.1	**0.0663**	0.022	32.5	ug/L	16	Standard
>	**In**	115	588246.1	1.2				ug/L	611117	Standard
	Sn	118	1550.1	10.3	**0.1702**	0.025	14.7	ug/L	518	Standard
	Sb	123	1125.9	6.3	**0.2350**	0.013	5.6	ug/L	127	Standard
	Ba	135	5627.4	4.0	**2.8066**	0.107	3.8	ug/L	45	Standard
	Ce	140	5279.3	1.4				ug/L	298	Standard
>	**Tb**	159	929768.7	1.5				ug/L	977942	Standard
	Ho	165	90.0	24.2				ug/L	12	Standard
	Tl	203	125.7	60.8	**0.0104**	0.009	86.4	ug/L	31	Standard
	Tl	205	76.7	65.3	**0.0084**	0.007	82.2	ug/L	20	Standard
	Pb	206	9321.2	0.6	**1.7003**	0.038	2.2	ug/L	286	Standard
	Pb	207	7700.6	1.2	**1.5576**	0.043	2.8	ug/L	257	Standard
	Pb	208	32126.4	0.5	**1.6815**	0.027	1.6	ug/L	956	Standard
	U	238	95.0	19.2	**0.0106**	0.002	23.3	ug/L	17	Standard
>	**Bi**	209	544815.4	1.6				ug/L	557080	Standard

Sample ID: L1606139307
Report Date/Time: Tuesday, June 28, 2016 22:59:03
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	241.7	8.4	**0.2600**	0.024	9.1	mg/L	32	Standard
K	39	93.3	13.5	**0.1697**	0.044	25.8	mg/L	30	Standard
Ca	43	48.3	11.9	**1.8938**	0.874	46.1	mg/L	48	Standard
Fe	54	296.6	3.5	**0.1277**	0.009	7.4	mg/L	159	Standard
Fe	57	275.0	13.7	**0.2917**	0.134	46.0	mg/L	195	Standard
> **Sc-1**	45	40101.3	1.0				mg/L	39543	Standard
Cl	35	33371.4	1.5				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	12191.6	6.9				ug/L	1647	Standard
P	31	16642.5	2.2				ug/L	15231	Standard
S	34	2221.8	1.3				ug/L	1845	Standard
Sr	88	81.7	3.5				ug/L	70	Standard
C	12	180.0	14.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	78.1				mg/L	0	Standard
Dy	164	179.2	25.7				mg/L	12	Standard
Ho-1	165	90.0	24.2				mg/L	12	Standard
Er	166	86.7	26.6				mg/L	20	Standard
I	127	5902.8	5.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		88.668	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.644	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606139307
Report Date/Time: Tuesday, June 28, 2016 22:59:03
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.258
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.798
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606140101
Sample Date/Time: Tuesday, June 28, 2016 22:59:58
Number of Replicates: 3
Autosampler Position: 350
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81815.4	2.6				ug/L	81760	Standard
	Be	9	21.7	13.3	**0.0121**	0.003	22.9	ug/L	17	Standard
	Al	27	45785308.7	4.2	**296.7336**	5.244	1.8	ug/L	1410	Standard
	Sc	45	41065.6	3.8				ug/L	39543	Standard
	Ti	47	178.0	12.2	**0.5887**	0.089	15.1	ug/L	39	Standard
	V	51	35448.4	4.1	**5.1643**	0.153	3.0	ug/L	1415	Standard
	Cr	52	9787.8	5.7	**0.4235**	0.068	16.1	ug/L	6946	Standard
	Cr	53	17411.8	14.2	**20.5297**	2.734	13.3	ug/L	328	Standard
	Mn	55	165972.1	5.7	**26.8086**	1.220	4.6	ug/L	823	Standard
	Co	59	1056.4	5.8	**0.1270**	0.007	5.7	ug/L	252	Standard
	Ni	60	3143.3	7.3	**1.6522**	0.109	6.6	ug/L	192	Standard
	Cu	65	2057.8	4.7	**1.0496**	0.040	3.9	ug/L	141	Standard
	Zn	66	6024.5	5.7	**4.9599**	0.261	5.3	ug/L	491	Standard
>	**Ge**	72	489001.5	1.3				ug/L	515087	Standard
	As	75	1112.2	8.9	**1.0817**	0.080	7.4	ug/L	20	Standard
	Se	82	302.7	9.2	**2.6482**	0.222	8.4	ug/L	15	Standard
	Se-1	77	1245.1	18.8	**16.6637**	3.078	18.5	ug/L	48	Standard
>	**Ga**	71	58.3	55.1				mg/L	90	Standard
	Rb	85	2691.9	5.7				ug/L	23	Standard
	Y	89	432812.2	6.0				ug/L	448510	Standard
>	**Rh**	103	128.3	14.8				ug/L	15	Standard
	Mo	98	642.1	7.2	**0.2310**	0.011	4.8	ug/L	11	Standard
	Ag	107	89.7	20.6	**-0.0021**	0.003	162.7	ug/L	134	Standard
	Cd	111	183.2	13.1	**0.0876**	0.014	15.6	mg/L	7	Standard
	Cd	114	423.4	6.2	**0.0810**	0.006	6.9	ug/L	16	Standard
>	**In**	115	580206.4	3.0				ug/L	611117	Standard
	Sn	118	1115.0	4.7	**0.0968**	0.010	9.8	ug/L	518	Standard
	Sb	123	200.7	10.2	**0.0319**	0.003	10.9	ug/L	127	Standard
	Ba	135	44440.1	3.3	**22.6462**	0.082	0.4	ug/L	45	Standard
	Ce	140	446.7	12.1				ug/L	298	Standard
>	**Tb**	159	929235.2	2.2				ug/L	977942	Standard
	Ho	165	38.3	7.5				ug/L	12	Standard
	Tl	203	126.0	16.1	**0.0119**	0.003	21.7	ug/L	31	Standard
	Tl	205	103.3	17.0	**0.0135**	0.003	21.8	ug/L	20	Standard
	Pb	206	396.7	5.3	**0.0183**	0.002	11.9	ug/L	286	Standard
	Pb	207	341.7	4.5	**0.0202**	0.002	8.5	ug/L	257	Standard
	Pb	208	1505.0	2.0	**0.0286**	0.002	5.8	ug/L	956	Standard
	U	238	10687.1	4.8	**1.6615**	0.054	3.3	ug/L	17	Standard
>	**Bi**	209	499114.9	2.7				ug/L	557080	Standard

Sample ID: L1606140101
Report Date/Time: Tuesday, June 28, 2016 23:02:15
Page 1

Approved: June 29, 2016

Na	23	8.3	91.7	**52.4138**	47.655	90.9	mg/L	0	Standard
Mg	24	165387.3	4.5	**211.9511**	6.406	3.0	mg/L	32	Standard
K	39	600.0	13.6	**1.8174**	0.318	17.5	mg/L	30	Standard
Ca	43	581.7	9.4	**79.6243**	5.149	6.5	mg/L	48	Standard
Fe	54	126.1	20.1	**-0.0269**	0.024	89.8	mg/L	159	Standard
Fe	57	591.7	6.9	**1.2999**	0.139	10.7	mg/L	195	Standard
> **Sc-1**	45	41065.6	3.8				mg/L	39543	Standard
Cl	35	47499.7	3.6				ug/L	27642	Standard
Kr	83	0.7	173.2				ug/L	2	Standard
Br	81	60832.3	7.6				ug/L	1647	Standard
P	31	21663.8	6.1				ug/L	15231	Standard
S	34	2320.2	3.1				ug/L	1845	Standard
Sr	88	470.0	1.1				ug/L	70	Standard
C	12	286.7	5.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	31.1	36.4				mg/L	12	Standard
Ho-1	165	38.3	7.5				mg/L	12	Standard
Er	166	46.7	44.6				mg/L	20	Standard
I	127	79897.5	13.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		100.068	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.936	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606140101
Report Date/Time: Tuesday, June 28, 2016 23:02:15
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.942
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		89.595
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606140201
Sample Date/Time: Tuesday, June 28, 2016 23:03:09
Number of Replicates: 3
Autosampler Position: 351
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76605.0	3.1				ug/L	81760	Standard
	Be	9	21.7	13.3	**0.0135**	0.003	20.6	ug/L	17	Standard
	Al	27	26105502.1	1.1	**180.8194**	3.801	2.1	ug/L	1410	Standard
	Sc	45	39097.0	0.2				ug/L	39543	Standard
	Ti	47	562.0	2.4	**2.2655**	0.057	2.5	ug/L	39	Standard
	V	51	4742.4	3.6	**0.5271**	0.022	4.2	ug/L	1415	Standard
	Cr	52	10082.4	1.2	**0.5007**	0.019	3.8	ug/L	6946	Standard
	Cr	53	12313.3	5.4	**14.7485**	0.907	6.2	ug/L	328	Standard
	Mn	55	1295871.4	1.2	**214.8489**	3.945	1.8	ug/L	823	Standard
	Co	59	1407.4	1.1	**0.1841**	0.002	1.1	ug/L	252	Standard
	Ni	60	4305.3	1.8	**2.3535**	0.034	1.4	ug/L	192	Standard
	Cu	65	1020.0	5.2	**0.4867**	0.034	6.9	ug/L	141	Standard
	Zn	66	18219.9	0.4	**17.0511**	0.173	1.0	ug/L	491	Standard
>	**Ge**	72	478933.9	0.7				ug/L	515087	Standard
	As	75	17040.6	2.0	**16.3399**	0.219	1.3	ug/L	20	Standard
	Se	82	892.6	3.4	**8.2180**	0.229	2.8	ug/L	15	Standard
	Se-1	77	753.4	7.3	**10.0390**	0.842	8.4	ug/L	48	Standard
>	**Ga**	71	218.3	6.6				mg/L	90	Standard
	Rb	85	3558.8	1.6				ug/L	23	Standard
	Y	89	416573.3	1.7				ug/L	448510	Standard
>	**Rh**	103	118.3	21.3				ug/L	15	Standard
	Mo	98	260.0	1.1	**0.0877**	0.004	4.4	ug/L	11	Standard
	Ag	107	80.3	26.9	**-0.0029**	0.004	134.9	ug/L	134	Standard
	Cd	111	40.9	26.1	**0.0184**	0.006	33.2	mg/L	7	Standard
	Cd	114	102.6	27.3	**0.0171**	0.006	37.8	ug/L	16	Standard
>	**In**	115	551143.5	2.5				ug/L	611117	Standard
	Sn	118	835.0	12.0	**0.0548**	0.020	37.1	ug/L	518	Standard
	Sb	123	829.1	3.5	**0.1820**	0.004	2.4	ug/L	127	Standard
	Ba	135	332622.5	0.7	**178.6804**	3.714	2.1	ug/L	45	Standard
	Ce	140	6094.6	2.7				ug/L	298	Standard
>	**Tb**	159	884304.9	2.4				ug/L	977942	Standard
	Ho	165	73.3	20.8				ug/L	12	Standard
	Tl	203	115.3	18.9	**0.0110**	0.003	28.4	ug/L	31	Standard
	Tl	205	95.0	13.9	**0.0127**	0.002	15.2	ug/L	20	Standard
	Pb	206	48021.0	1.3	**10.1649**	0.107	1.1	ug/L	286	Standard
	Pb	207	40222.3	0.8	**9.4401**	0.093	1.0	ug/L	257	Standard
	Pb	208	168471.0	1.1	**10.2223**	0.060	0.6	ug/L	956	Standard
	U	238	714.0	7.4	**0.1118**	0.009	8.4	ug/L	17	Standard
>	**Bi**	209	483931.7	1.1				ug/L	557080	Standard

Sample ID: L1606140201
Report Date/Time: Tuesday, June 28, 2016 23:05:26
Page 1

Approved: June 29, 2016

Na	23	8.3	34.6	**54.1394**	18.829	34.8	mg/L	0	Standard
Mg	24	106065.1	2.2	**142.7401**	3.063	2.1	mg/L	32	Standard
K	39	148.3	8.5	**0.3655**	0.044	12.0	mg/L	30	Standard
Ca	43	363.3	11.2	**50.4710**	6.294	12.5	mg/L	48	Standard
Fe	54	1618.9	1.4	**1.3448**	0.023	1.7	mg/L	159	Standard
Fe	57	835.0	5.7	**2.2263**	0.157	7.1	mg/L	195	Standard
> **Sc-1**	45	39097.0	0.2				mg/L	39543	Standard
Cl	35	36261.3	1.4				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	214333.6	4.1				ug/L	1647	Standard
P	31	21892.5	1.8				ug/L	15231	Standard
S	34	2305.2	3.8				ug/L	1845	Standard
Sr	88	363.3	12.3				ug/L	70	Standard
C	12	390.0	0.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	133.0	23.5				mg/L	12	Standard
Ho-1	165	73.3	20.8				mg/L	12	Standard
Er	166	76.7	66.9				mg/L	20	Standard
I	127	398539.7	6.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.695	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.981	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606140201
Report Date/Time: Tuesday, June 28, 2016 23:05:26
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		90.186
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		86.869
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606140201
Report Date/Time: Tuesday, June 28, 2016 23:05:26
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606140401
Sample Date/Time: Tuesday, June 28, 2016 23:06:21
Number of Replicates: 3
Autosampler Position: 352
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74993.5	0.5				ug/L	81760	Standard
	Be	9	21.7	13.3	**0.0141**	0.003	22.5	ug/L	17	Standard
	Al	27	30278310.9	1.4	**214.1402**	2.917	1.4	ug/L	1410	Standard
	Sc	45	39486.4	2.9				ug/L	39543	Standard
	Ti	47	193.3	4.5	**0.6334**	0.048	7.5	ug/L	39	Standard
	V	51	22966.6	2.4	**3.1851**	0.035	1.1	ug/L	1415	Standard
	Cr	52	11319.6	0.7	**0.6125**	0.031	5.1	ug/L	6946	Standard
	Cr	53	1811.8	4.8	**1.7776**	0.107	6.0	ug/L	328	Standard
	Mn	55	8657.1	1.2	**1.2226**	0.034	2.8	ug/L	823	Standard
	Co	59	565.7	7.8	**0.0513**	0.007	13.0	ug/L	252	Standard
	Ni	60	1945.8	2.7	**0.9557**	0.026	2.7	ug/L	192	Standard
	Cu	65	60353.0	0.9	**32.5191**	0.515	1.6	ug/L	141	Standard
	Zn	66	48175.8	1.3	**44.3004**	1.008	2.3	ug/L	491	Standard
>	**Ge**	72	501772.7	1.4				ug/L	515087	Standard
	As	75	367.9	8.8	**0.3760**	0.025	6.7	ug/L	20	Standard
	Se	82	110.4	13.6	**0.8639**	0.120	13.9	ug/L	15	Standard
	Se-1	77	129.0	8.1	**1.0381**	0.123	11.8	ug/L	48	Standard
>	**Ga**	71	68.3	27.7				mg/L	90	Standard
	Rb	85	1198.4	2.3				ug/L	23	Standard
	Y	89	417957.0	1.4				ug/L	448510	Standard
>	**Rh**	103	78.3	35.2				ug/L	15	Standard
	Mo	98	388.3	6.2	**0.1355**	0.009	6.9	ug/L	11	Standard
	Ag	107	94.3	4.8	**-0.0011**	0.001	61.6	ug/L	134	Standard
	Cd	111	42.7	19.5	**0.0186**	0.004	22.3	mg/L	7	Standard
	Cd	114	106.8	24.4	**0.0172**	0.005	31.6	ug/L	16	Standard
>	**In**	115	568064.8	0.3				ug/L	611117	Standard
	Sn	118	741.7	8.4	**0.0330**	0.012	35.5	ug/L	518	Standard
	Sb	123	163.0	3.3	**0.0243**	0.001	4.6	ug/L	127	Standard
	Ba	135	87619.3	0.7	**45.6318**	0.199	0.4	ug/L	45	Standard
	Ce	140	530.0	5.7				ug/L	298	Standard
>	**Tb**	159	902463.1	0.6				ug/L	977942	Standard
	Ho	165	31.7	24.1				ug/L	12	Standard
	Tl	203	89.3	3.6	**0.0068**	0.000	7.0	ug/L	31	Standard
	Tl	205	80.0	18.8	**0.0096**	0.002	23.2	ug/L	20	Standard
	Pb	206	4637.4	0.5	**0.8697**	0.009	1.0	ug/L	286	Standard
	Pb	207	3892.8	1.5	**0.8106**	0.017	2.1	ug/L	257	Standard
	Pb	208	15697.9	1.6	**0.8448**	0.009	1.1	ug/L	956	Standard
	U	238	6211.6	0.5	**0.9400**	0.010	1.1	ug/L	17	Standard
>	**Bi**	209	512163.0	0.7				ug/L	557080	Standard

Sample ID: L1606140401
Report Date/Time: Tuesday, June 28, 2016 23:08:38
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.8762**	18.829	173.1	mg/L	0	Standard
Mg	24	25888.8	0.8	**34.4779**	1.238	3.6	mg/L	32	Standard
K	39	206.7	17.0	**0.5565**	0.104	18.6	mg/L	30	Standard
Ca	43	238.3	20.2	**30.9565**	7.726	25.0	mg/L	48	Standard
Fe	54	118.7	26.0	**-0.0297**	0.026	88.1	mg/L	159	Standard
Fe	57	376.7	12.0	**0.6523**	0.186	28.5	mg/L	195	Standard
> **Sc-1**	45	39486.4	2.9				mg/L	39543	Standard
Cl	35	38202.8	1.2				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	11808.0	7.7				ug/L	1647	Standard
P	31	21151.4	0.3				ug/L	15231	Standard
S	34	2466.9	2.1				ug/L	1845	Standard
Sr	88	291.7	10.5				ug/L	70	Standard
C	12	196.7	43.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	28.2	2.0				mg/L	12	Standard
Ho-1	165	31.7	24.1				mg/L	12	Standard
Er	166	36.7	31.5				mg/L	20	Standard
I	127	34317.5	11.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.724	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.415	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606140401
Report Date/Time: Tuesday, June 28, 2016 23:08:38
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		92.955
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		91.937
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606140401
Report Date/Time: Tuesday, June 28, 2016 23:08:38
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606140801
Sample Date/Time: Tuesday, June 28, 2016 23:09:32
Number of Replicates: 3
Autosampler Position: 353
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78174.6	2.2				ug/L	81760	Standard
	Be	9	35.0	24.7	**0.0270**	0.008	30.6	ug/L	17	Standard
	Al	27	18286645.8	1.2	**124.0948**	2.663	2.1	ug/L	1410	Standard
	Sc	45	40988.7	1.5				ug/L	39543	Standard
	Ti	47	225.3	13.4	**0.7985**	0.121	15.1	ug/L	39	Standard
	V	51	51258.6	3.2	**7.6390**	0.153	2.0	ug/L	1415	Standard
	Cr	52	11486.4	2.6	**0.6992**	0.069	9.9	ug/L	6946	Standard
	Cr	53	8145.5	5.8	**9.5287**	0.474	5.0	ug/L	328	Standard
	Mn	55	98520.0	1.7	**16.0208**	0.120	0.7	ug/L	823	Standard
	Co	59	769.4	2.9	**0.0851**	0.002	2.7	ug/L	252	Standard
	Ni	60	1615.4	2.4	**0.8080**	0.032	4.0	ug/L	192	Standard
	Cu	65	1131.0	3.0	**0.5425**	0.014	2.7	ug/L	141	Standard
	Zn	66	4403.0	3.4	**3.4460**	0.103	3.0	ug/L	491	Standard
>	**Ge**	72	484103.1	1.4				ug/L	515087	Standard
	As	75	1970.8	6.0	**1.9047**	0.087	4.5	ug/L	20	Standard
	Se	82	191.5	8.8	**1.6489**	0.132	8.0	ug/L	15	Standard
	Se-1	77	529.7	10.7	**6.7555**	0.697	10.3	ug/L	48	Standard
>	**Ga**	71	85.0	31.1				mg/L	90	Standard
	Rb	85	2531.9	3.1				ug/L	23	Standard
	Y	89	421592.5	4.0				ug/L	448510	Standard
>	**Rh**	103	66.7	30.3				ug/L	15	Standard
	Mo	98	353.3	5.0	**0.1209**	0.007	5.4	ug/L	11	Standard
	Ag	107	105.0	57.1	**0.0006**	0.010	1741.0	ug/L	134	Standard
	Cd	111	58.7	7.1	**0.0265**	0.001	5.4	mg/L	7	Standard
	Cd	114	139.6	12.3	**0.0238**	0.003	13.4	ug/L	16	Standard
>	**In**	115	570844.6	2.8				ug/L	611117	Standard
	Sn	118	760.0	5.7	**0.0358**	0.009	25.3	ug/L	518	Standard
	Sb	123	136.8	41.6	**0.0182**	0.013	71.7	ug/L	127	Standard
	Ba	135	27343.0	2.9	**14.1534**	0.097	0.7	ug/L	45	Standard
	Ce	140	1776.8	2.1				ug/L	298	Standard
>	**Tb**	159	907261.5	0.9				ug/L	977942	Standard
	Ho	165	38.3	32.8				ug/L	12	Standard
	Tl	203	160.7	51.8	**0.0162**	0.011	66.8	ug/L	31	Standard
	Tl	205	123.3	9.4	**0.0162**	0.002	11.0	ug/L	20	Standard
	Pb	206	469.0	16.1	**0.0318**	0.015	47.2	ug/L	286	Standard
	Pb	207	405.3	12.5	**0.0333**	0.011	33.4	ug/L	257	Standard
	Pb	208	1576.7	3.7	**0.0314**	0.003	10.7	ug/L	956	Standard
	U	238	14465.9	3.7	**2.2164**	0.080	3.6	ug/L	17	Standard
>	**Bi**	209	506765.8	0.4				ug/L	557080	Standard

Sample ID: L1606140801
Report Date/Time: Tuesday, June 28, 2016 23:11:49
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**20.4816**	17.733	86.6	mg/L	0	Standard
Mg	24	76429.0	2.1	**98.0879**	0.872	0.9	mg/L	32	Standard
K	39	103.3	20.1	**0.1958**	0.071	36.2	mg/L	30	Standard
Ca	43	268.3	18.7	**34.0467**	7.923	23.3	mg/L	48	Standard
Fe	54	210.5	5.9	**0.0466**	0.009	18.4	mg/L	159	Standard
Fe	57	346.7	13.1	**0.5032**	0.132	26.3	mg/L	195	Standard
Sc-1	45	40988.7	1.5				mg/L	39543	Standard
Cl	35	41040.2	0.8				ug/L	27642	Standard
Kr	83	2.0	173.2				ug/L	2	Standard
Br	81	20580.7	5.3				ug/L	1647	Standard
P	31	21832.4	1.0				ug/L	15231	Standard
S	34	2461.9	6.1				ug/L	1845	Standard
Sr	88	228.3	13.2				ug/L	70	Standard
C	12	196.7	34.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	38.1	77.6				mg/L	12	Standard
Ho-1	165	38.3	32.8				mg/L	12	Standard
Er	166	40.0	90.1				mg/L	20	Standard
I	127	105209.9	5.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		95.615	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.985	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606140801
Report Date/Time: Tuesday, June 28, 2016 23:11:49
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	93.410
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	90.968
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606140901
Sample Date/Time: Tuesday, June 28, 2016 23:12:44
Number of Replicates: 3
Autosampler Position: 354
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78595.1	3.1				ug/L	81760	Standard
	Be	9	26.7	57.3	**0.0179**	0.015	85.3	ug/L	17	Standard
	Al	27	20167615.6	1.1	**136.1889**	4.854	3.6	ug/L	1410	Standard
	Sc	45	39818.9	1.9				ug/L	39543	Standard
	Ti	47	311.7	10.4	**1.1846**	0.135	11.4	ug/L	39	Standard
	V	51	55787.4	0.6	**8.4391**	0.018	0.2	ug/L	1415	Standard
	Cr	52	16179.6	1.9	**1.4466**	0.047	3.2	ug/L	6946	Standard
	Cr	53	12833.8	2.8	**15.4070**	0.434	2.8	ug/L	328	Standard
	Mn	55	13187.4	1.1	**2.0427**	0.028	1.4	ug/L	823	Standard
	Co	59	670.0	4.7	**0.0714**	0.005	7.2	ug/L	252	Standard
	Ni	60	2540.2	2.3	**1.3479**	0.032	2.4	ug/L	192	Standard
	Cu	65	810.4	1.3	**0.3686**	0.005	1.5	ug/L	141	Standard
	Zn	66	2808.3	3.3	**1.9320**	0.083	4.3	ug/L	491	Standard
>	**Ge**	72	478186.0	0.4				ug/L	515087	Standard
	As	75	1572.1	0.9	**1.5465**	0.010	0.6	ug/L	20	Standard
	Se	82	338.5	3.3	**3.0475**	0.114	3.7	ug/L	15	Standard
	Se-1	77	792.0	6.7	**10.6055**	0.757	7.1	ug/L	48	Standard
>	**Ga**	71	71.7	10.7				mg/L	90	Standard
	Rb	85	3450.4	2.1				ug/L	23	Standard
	Y	89	421286.0	4.0				ug/L	448510	Standard
>	**Rh**	103	76.7	10.0				ug/L	15	Standard
	Mo	98	1179.3	2.2	**0.4594**	0.016	3.6	ug/L	11	Standard
	Ag	107	70.0	6.5	**-0.0048**	0.001	13.6	ug/L	134	Standard
	Cd	111	26.3	13.5	**0.0106**	0.002	18.0	mg/L	7	Standard
	Cd	114	71.5	14.0	**0.0102**	0.002	20.9	ug/L	16	Standard
>	**In**	115	557486.7	1.6				ug/L	611117	Standard
	Sn	118	753.4	2.1	**0.0377**	0.003	7.4	ug/L	518	Standard
	Sb	123	140.1	13.9	**0.0197**	0.005	23.1	ug/L	127	Standard
	Ba	135	30044.1	2.0	**15.9271**	0.116	0.7	ug/L	45	Standard
	Ce	140	1501.7	8.9				ug/L	298	Standard
>	**Tb**	159	887600.8	1.1				ug/L	977942	Standard
	Ho	165	38.3	27.2				ug/L	12	Standard
	Tl	203	136.3	15.5	**0.0135**	0.003	21.6	ug/L	31	Standard
	Tl	205	116.7	8.9	**0.0158**	0.002	10.2	ug/L	20	Standard
	Pb	206	477.7	6.1	**0.0364**	0.006	16.6	ug/L	286	Standard
	Pb	207	398.0	3.8	**0.0344**	0.003	8.7	ug/L	257	Standard
	Pb	208	1707.0	2.5	**0.0420**	0.002	4.6	ug/L	956	Standard
	U	238	23354.0	0.6	**3.6864**	0.026	0.7	ug/L	17	Standard
>	**Bi**	209	492176.6	0.6				ug/L	557080	Standard

Sample ID: L1606140901
Report Date/Time: Tuesday, June 28, 2016 23:15:01
Page 1

Approved: June 29, 2016

Na	23	8.3	124.9	**53.9308**	68.028	126.1	mg/L	0	Standard
Mg	24	78221.4	0.2	**103.3688**	1.796	1.7	mg/L	32	Standard
K	39	181.7	9.7	**0.4684**	0.061	12.9	mg/L	30	Standard
Ca	43	295.0	5.9	**39.1939**	3.495	8.9	mg/L	48	Standard
Fe	54	173.9	5.6	**0.0192**	0.007	35.0	mg/L	159	Standard
Fe	57	403.3	17.1	**0.7302**	0.251	34.4	mg/L	195	Standard
> **Sc-1**	45	39818.9	1.9				mg/L	39543	Standard
Cl	35	38063.1	1.8				ug/L	27642	Standard
Kr	83	1.3	173.2				ug/L	2	Standard
Br	81	39401.2	2.9				ug/L	1647	Standard
P	31	23753.6	2.5				ug/L	15231	Standard
S	34	2233.5	3.9				ug/L	1845	Standard
Sr	88	295.0	6.1				ug/L	70	Standard
C	12	220.0	31.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	222.1	127.8				mg/L	12	Standard
Ho-1	165	38.3	27.2				mg/L	12	Standard
Er	166	26.7	78.1				mg/L	20	Standard
I	127	56189.0	1.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		96.129	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.836	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606140901
Report Date/Time: Tuesday, June 28, 2016 23:15:01
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		91.224
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		88.349
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606140901
Report Date/Time: Tuesday, June 28, 2016 23:15:01
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 23:15:56
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	73471.1	1.2				ug/L	81760	Standard
	Be	9	47618.1	3.0	**53.3659**	2.229	4.2	ug/L	17	Standard
	Al	27	7287296.4	1.0	**52.5966**	0.249	0.5	ug/L	1410	Standard
	Sc	45	40183.2	3.0				ug/L	39543	Standard
	Ti	47	24656.7	0.7	**99.9368**	1.445	1.4	ug/L	39	Standard
	V	51	342959.6	1.5	**49.6067**	0.741	1.5	ug/L	1415	Standard
	Cr	52	350575.7	1.6	**49.7880**	0.923	1.9	ug/L	6946	Standard
	Cr	53	44859.7	1.2	**51.2482**	0.358	0.7	ug/L	328	Standard
	Mn	55	313688.7	1.9	**48.6980**	0.765	1.6	ug/L	823	Standard
	Co	59	336351.1	1.2	**48.2963**	0.507	1.1	ug/L	252	Standard
	Ni	60	92893.3	2.1	**49.6893**	0.993	2.0	ug/L	192	Standard
	Cu	65	92858.0	0.8	**49.2454**	0.445	0.9	ug/L	141	Standard
	Zn	66	54854.1	1.5	**49.6985**	0.961	1.9	ug/L	491	Standard
>	Ge	72	510232.1	0.7				ug/L	515087	Standard
	As	75	53972.2	1.6	**48.5043**	0.935	1.9	ug/L	20	Standard
	Se	82	5467.9	0.8	**47.8296**	0.724	1.5	ug/L	15	Standard
	Se-1	77	3829.5	1.2	**50.5966**	0.465	0.9	ug/L	48	Standard
>	Ga	71	55.0	32.8				mg/L	90	Standard
	Rb	85	740.0	22.2				ug/L	23	Standard
	Y	89	417471.8	0.9				ug/L	448510	Standard
>	Rh	103	28.3	40.8				ug/L	15	Standard
	Mo	98	255774.9	1.1	**100.4324**	1.068	1.1	ug/L	11	Standard
	Ag	107	314263.4	0.6	**49.6826**	0.310	0.6	ug/L	134	Standard
	Cd	111	99244.0	2.0	**49.2800**	0.642	1.3	mg/L	7	Standard
	Cd	114	244609.6	0.7	**49.9721**	0.117	0.2	ug/L	16	Standard
>	In	115	575695.1	0.7				ug/L	611117	Standard
	Sn	118	277518.7	1.2	**49.7196**	0.393	0.8	ug/L	518	Standard
	Sb	123	223405.3	1.0	**50.2498**	0.315	0.6	ug/L	127	Standard
	Ba	135	95362.2	0.9	**49.0082**	0.194	0.4	ug/L	45	Standard
	Ce	140	66.7	18.9				ug/L	298	Standard
>	Tb	159	900469.8	1.2				ug/L	977942	Standard
	Ho	165	0.0					ug/L	12	Standard
	Tl	203	399704.8	0.8	**50.1537**	0.324	0.6	ug/L	31	Standard
	Tl	205	354566.9	2.1	**51.2307**	0.689	1.3	ug/L	20	Standard
	Pb	206	258134.8	1.1	**50.4360**	0.346	0.7	ug/L	286	Standard
	Pb	207	234508.0	1.0	**50.8014**	0.140	0.3	ug/L	257	Standard
	Pb	208	934488.0	0.8	**52.3272**	0.390	0.7	ug/L	956	Standard
	U	238	348175.4	0.3	**51.3787**	0.514	1.0	ug/L	17	Standard
>	Bi	209	526902.0	1.3				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:18:13
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**32.0442**	31.868	99.5	mg/L	0	Standard
Mg	24	3947.2	3.3	**5.1157**	0.206	4.0	mg/L	32	Standard
K	39	1291.7	6.0	**4.1569**	0.313	7.5	mg/L	30	Standard
Ca	43	78.3	13.3	**6.3383**	1.191	18.8	mg/L	48	Standard
Fe	54	5567.6	3.6	**4.8270**	0.308	6.4	mg/L	159	Standard
Fe	57	1653.4	2.1	**4.8734**	0.265	5.4	mg/L	195	Standard
> **Sc-1**	45	40183.2	3.0				mg/L	39543	Standard
Cl	35	36099.6	2.2				ug/L	27642	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	2486.9	19.4				ug/L	1647	Standard
P	31	18199.3	2.9				ug/L	15231	Standard
S	34	2558.5	2.4				ug/L	1845	Standard
Sr	88	88.3	13.1				ug/L	70	Standard
C	12	123.3	18.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	2.5	205.9				mg/L	12	Standard
Ho-1	165	0.0					mg/L	12	Standard
Er	166	16.7	69.3				mg/L	20	Standard
I	127	3410.4	24.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	106.732		
	Al	27	105.193		
	Sc	45			
	Ti	47	99.937		
	V	51	99.213		
	Cr	52	99.576		
	Cr	53			
	Mn	55	97.396		
	Co	59	96.593		
	Ni	60	99.379		
	Cu	65	98.491		
	Zn	66	99.397		
>	Ge	72		99.057	
	As	75	97.009		
	Se	82	95.659		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:18:13
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	100.432		
Ag	107	99.365		
Cd	111	98.560		
Cd	114			
In	115			94.204
Sn	118	99.439		
Sb	123	100.500		
Ba	135	98.016		
Ce	140			
Tb	159			
Ho	165			
Tl	203	100.307		
Tl	205			
Pb	206			
Pb	207			
Pb	208	104.654		
U	238	102.757		
Bi	209			94.583
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:18:13
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 23:19:08
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74688.6	0.5				ug/L	81760	Standard
	Be	9	23.3	68.9	**0.0159**	0.018	110.7	ug/L	17	Standard
	Al	27	6963.5	51.4	**0.0350**	0.025	72.0	ug/L	1410	Standard
	Sc	45	39526.5	2.8				ug/L	39543	Standard
	Ti	47	29.3	2.0	**-0.0430**	0.003	6.9	ug/L	39	Standard
	V	51	1022.6	3.0	**-0.0543**	0.006	11.1	ug/L	1415	Standard
	Cr	52	6275.6	1.6	**-0.1243**	0.005	4.1	ug/L	6946	Standard
	Cr	53	655.0	2.0	**0.4308**	0.021	5.0	ug/L	328	Standard
	Mn	55	708.0	10.7	**-0.0355**	0.011	31.8	ug/L	823	Standard
	Co	59	254.0	9.4	**0.0060**	0.003	53.0	ug/L	252	Standard
	Ni	60	190.7	1.6	**-0.0002**	0.001	475.8	ug/L	192	Standard
	Cu	65	144.3	16.1	**-0.0123**	0.012	100.6	ug/L	141	Standard
	Zn	66	441.0	4.3	**-0.4114**	0.015	3.7	ug/L	491	Standard
>	**Ge**	72	497792.7	1.0				ug/L	515087	Standard
	As	75	-15.2	258.6	**0.0262**	0.036	138.8	ug/L	20	Standard
	Se	82	23.5	16.0	**0.0916**	0.033	35.7	ug/L	15	Standard
	Se-1	77	63.7	3.3	**0.1558**	0.037	23.6	ug/L	48	Standard
>	**Ga**	71	50.0	34.6				mg/L	90	Standard
	Rb	85	33.3	56.8				ug/L	23	Standard
	Y	89	425942.8	1.0				ug/L	448510	Standard
>	**Rh**	103	6.7	43.3				ug/L	15	Standard
	Mo	98	281.0	37.4	**0.0876**	0.040	45.9	ug/L	11	Standard
	Ag	107	137.7	21.7	**0.0048**	0.005	95.8	ug/L	134	Standard
	Cd	111	15.2	29.8	**0.0044**	0.002	50.1	mg/L	7	Standard
	Cd	114	43.1	89.6	**0.0036**	0.008	212.7	ug/L	16	Standard
>	**In**	115	595727.8	0.7				ug/L	611117	Standard
	Sn	118	753.4	17.4	**0.0287**	0.023	78.8	ug/L	518	Standard
	Sb	123	441.8	17.1	**0.0832**	0.016	19.8	ug/L	127	Standard
	Ba	135	56.0	54.4	**0.0027**	0.015	556.5	ug/L	45	Standard
	Ce	140	23.3	12.4				ug/L	298	Standard
>	**Tb**	159	931120.0	0.7				ug/L	977942	Standard
	Ho	165	10.0	86.6				ug/L	12	Standard
	Tl	203	88.3	51.6	**0.0060**	0.006	92.0	ug/L	31	Standard
	Tl	205	63.3	63.8	**0.0066**	0.006	85.2	ug/L	20	Standard
	Pb	206	308.0	15.6	**-0.0049**	0.009	177.6	ug/L	286	Standard
	Pb	207	259.7	11.6	**-0.0032**	0.006	187.4	ug/L	257	Standard
	Pb	208	1046.3	7.8	**-0.0033**	0.004	126.3	ug/L	956	Standard
	U	238	64.3	43.1	**0.0063**	0.004	62.8	ug/L	17	Standard
>	**Bi**	209	540861.2	0.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 23:21:25
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.5148**	18.203	173.1	mg/L	0	Standard
Mg	24	55.0	41.7	**0.0162**	0.030	187.5	mg/L	32	Standard
K	39	25.0	20.0	**-0.0566**	0.019	34.0	mg/L	30	Standard
Ca	43	33.3	34.6	**-0.2594**	1.827	704.2	mg/L	48	Standard
Fe	54	134.2	10.1	**-0.0157**	0.009	57.1	mg/L	159	Standard
Fe	57	241.7	9.3	**0.1932**	0.098	50.6	mg/L	195	Standard
Sc-1	45	39526.5	2.8				mg/L	39543	Standard
Cl	35	35813.0	1.8				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1830.1	10.1				ug/L	1647	Standard
P	31	18830.0	1.9				ug/L	15231	Standard
S	34	2625.2	3.0				ug/L	1845	Standard
Sr	88	65.0	33.5				ug/L	70	Standard
C	12	80.0	21.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	19.2	93.8				mg/L	12	Standard
Ho-1	165	10.0	86.6				mg/L	12	Standard
Er	166	16.7	91.7				mg/L	20	Standard
I	127	2391.9	12.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.642	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 23:21:25
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.482
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.089
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 23:21:25
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Tuesday, June 28, 2016 23:22:21
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70209.0	1.7				ug/L	81760	Standard
	Be	9	180.0	16.9	**0.2017**	0.039	19.3	ug/L	17	Standard
	Al	27	5133.2	109.4	**0.0240**	0.041	173.1	ug/L	1410	Standard
	Sc	45	36931.6	2.6				ug/L	39543	Standard
	Ti	47	20.7	36.3	**-0.0758**	0.033	44.0	ug/L	39	Standard
	V	51	3515.4	3.3	**0.3352**	0.014	4.2	ug/L	1415	Standard
	Cr	52	11116.7	0.6	**0.6541**	0.028	4.2	ug/L	6946	Standard
	Cr	53	1220.0	4.0	**1.1474**	0.049	4.3	ug/L	328	Standard
	Mn	55	3585.8	2.9	**0.4442**	0.023	5.1	ug/L	823	Standard
	Co	59	2687.2	1.7	**0.3785**	0.013	3.3	ug/L	252	Standard
	Ni	60	2873.6	1.1	**1.5299**	0.031	2.0	ug/L	192	Standard
	Cu	65	1483.4	2.9	**0.7455**	0.015	2.0	ug/L	141	Standard
	Zn	66	6969.9	1.2	**5.9869**	0.185	3.1	ug/L	491	Standard
>	**Ge**	72	480794.8	1.6				ug/L	515087	Standard
	As	75	396.5	4.3	**0.4183**	0.021	5.0	ug/L	20	Standard
	Se	82	61.9	5.1	**0.4561**	0.020	4.4	ug/L	15	Standard
	Se-1	77	78.3	14.0	**0.3969**	0.172	43.4	ug/L	48	Standard
>	**Ga**	71	46.7	32.7				mg/L	90	Standard
	Rb	85	21.7	13.3				ug/L	23	Standard
	Y	89	400775.4	1.5				ug/L	448510	Standard
>	**Rh**	103	13.3	21.7				ug/L	15	Standard
	Mo	98	70.5	1.1	**0.0095**	0.000	1.9	ug/L	11	Standard
	Ag	107	2377.2	2.1	**0.3719**	0.012	3.3	ug/L	134	Standard
	Cd	111	440.9	3.8	**0.2230**	0.005	2.1	mg/L	7	Standard
	Cd	114	1065.8	11.0	**0.2200**	0.026	11.9	ug/L	16	Standard
>	**In**	115	557761.4	1.6				ug/L	611117	Standard
	Sn	118	468.3	18.8	**-0.0150**	0.017	116.5	ug/L	518	Standard
	Sb	123	1709.9	1.5	**0.3843**	0.009	2.4	ug/L	127	Standard
	Ba	135	1358.4	3.9	**0.6957**	0.022	3.2	ug/L	45	Standard
	Ce	140	15.0	33.3				ug/L	298	Standard
>	**Tb**	159	873882.2	1.3				ug/L	977942	Standard
	Ho	165	8.3	124.9				ug/L	12	Standard
	Tl	203	633.3	2.0	**0.0774**	0.002	2.5	ug/L	31	Standard
	Tl	205	590.0	17.7	**0.0858**	0.015	17.2	ug/L	20	Standard
	Pb	206	1230.0	0.6	**0.1853**	0.002	0.8	ug/L	286	Standard
	Pb	207	1060.4	4.9	**0.1798**	0.012	6.4	ug/L	257	Standard
	Pb	208	4496.6	3.9	**0.2002**	0.009	4.3	ug/L	956	Standard
	U	238	2637.6	2.9	**0.3996**	0.008	2.0	ug/L	17	Standard
>	**Bi**	209	509391.4	0.9				ug/L	557080	Standard

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 23:24:38
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	38.3	30.1	**-0.0027**	0.016	586.9	mg/L	32	Standard
K	39	28.3	44.4	**-0.0396**	0.043	109.0	mg/L	30	Standard
Ca	43	26.7	43.3	**-1.0379**	1.747	168.3	mg/L	48	Standard
Fe	54	112.3	22.9	**-0.0284**	0.024	85.0	mg/L	159	Standard
Fe	57	253.3	17.2	**0.2902**	0.150	51.8	mg/L	195	Standard
> **Sc-1**	45	36931.6	2.6				mg/L	39543	Standard
Cl	35	34628.2	2.2				ug/L	27642	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	1663.4	2.8				ug/L	1647	Standard
P	31	13167.4	6.3				ug/L	15231	Standard
S	34	2550.2	3.7				ug/L	1845	Standard
Sr	88	81.7	12.7				ug/L	70	Standard
C	12	96.7	15.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	22.7	104.3				mg/L	12	Standard
Ho-1	165	8.3	124.9				mg/L	12	Standard
Er	166	13.3	173.2				mg/L	20	Standard
I	127	1815.1	19.4				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	100.854		
	Al	27			
	Sc	45			
	Ti	47			
	V	51	83.798		
	Cr	52	81.760		
	Cr	53			
	Mn	55	88.848		
	Co	59	94.629		
	Ni	60	95.622		
	Cu	65	93.183		
	Zn	66	95.791		
>	Ge	72		93.342	
	As	75	104.574		
	Se	82	114.027		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 23:24:38
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	92.977		
Cd	111	92.929		
Cd	114			
> In	115			91.269
Sn	118			
Sb	123	96.075		
Ba	135	92.766		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	96.789		
Tl	205			
Pb	206			
Pb	207			
Pb	208	100.122		
U	238	99.892		
> Bi	209			91.439
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: PBS M4 WG574306-02
Sample Date/Time: Tuesday, June 28, 2016 23:25:34
Number of Replicates: 3
Autosampler Position: 401
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73219.9	2.0				ug/L	81760	Standard
	Be	9	21.7	13.3	**0.0146**	0.003	19.0	ug/L	17	Standard
	Al	27	8870.9	1.6	**0.0499**	0.001	2.1	ug/L	1410	Standard
	Sc	45	39242.4	3.0				ug/L	39543	Standard
	Ti	47	26.0	21.4	**-0.0565**	0.021	37.8	ug/L	39	Standard
	V	51	1183.1	4.4	**-0.0293**	0.008	27.7	ug/L	1415	Standard
	Cr	52	7819.4	0.5	**0.1132**	0.035	31.1	ug/L	6946	Standard
	Cr	53	706.7	5.7	**0.4968**	0.043	8.6	ug/L	328	Standard
	Mn	55	910.4	1.2	**-0.0022**	0.005	247.3	ug/L	823	Standard
	Co	59	210.7	7.4	**-0.0001**	0.003	2281.1	ug/L	252	Standard
	Ni	60	270.7	10.1	**0.0449**	0.018	40.4	ug/L	192	Standard
	Cu	65	197.3	7.2	**0.0174**	0.009	50.7	ug/L	141	Standard
	Zn	66	1637.4	3.9	**0.7282**	0.056	7.7	ug/L	491	Standard
>	**Ge**	72	494675.2	2.7				ug/L	515087	Standard
	As	75	-22.9	72.5	**0.0188**	0.016	85.6	ug/L	20	Standard
	Se	82	19.6	24.9	**0.0572**	0.041	70.8	ug/L	15	Standard
	Se-1	77	57.0	14.0	**0.0674**	0.090	134.0	ug/L	48	Standard
>	**Ga**	71	43.3	17.6				mg/L	90	Standard
	Rb	85	43.3	35.3				ug/L	23	Standard
	Y	89	424600.8	0.9				ug/L	448510	Standard
>	**Rh**	103	15.0	66.7				ug/L	15	Standard
	Mo	98	68.5	24.2	**0.0070**	0.006	90.8	ug/L	11	Standard
	Ag	107	79.3	10.7	**-0.0041**	0.001	32.3	ug/L	134	Standard
	Cd	111	8.5	44.3	**0.0012**	0.002	151.3	mg/L	7	Standard
	Cd	114	29.3	97.6	**0.0009**	0.006	647.6	ug/L	16	Standard
>	**In**	115	594639.0	0.2				ug/L	611117	Standard
	Sn	118	648.3	4.2	**0.0107**	0.005	42.7	ug/L	518	Standard
	Sb	123	146.0	12.4	**0.0190**	0.004	20.5	ug/L	127	Standard
	Ba	135	94.3	2.4	**0.0219**	0.001	5.5	ug/L	45	Standard
	Ce	140	41.7	27.7				ug/L	298	Standard
>	**Tb**	159	926166.4	1.8				ug/L	977942	Standard
	Ho	165	11.7	24.7				ug/L	12	Standard
	Tl	203	20.7	43.9	**-0.0022**	0.001	49.4	ug/L	31	Standard
	Tl	205	15.0	33.3	**-0.0001**	0.001	484.0	ug/L	20	Standard
	Pb	206	313.0	7.9	**-0.0032**	0.004	138.6	ug/L	286	Standard
	Pb	207	286.3	1.1	**0.0033**	0.001	36.5	ug/L	257	Standard
	Pb	208	1144.7	1.9	**0.0029**	0.001	51.3	ug/L	956	Standard
	U	238	10.7	32.9	**-0.0014**	0.001	35.7	ug/L	17	Standard
>	**Bi**	209	533593.4	0.9				ug/L	557080	Standard

Sample ID: PBS M4 WG574306-02
Report Date/Time: Tuesday, June 28, 2016 23:27:51
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	31.7	39.7	**-0.0149**	0.016	104.8	mg/L	32	Standard
K	39	41.7	69.3	**0.0013**	0.101	7981.1	mg/L	30	Standard
Ca	43	18.3	15.7	**-2.5418**	0.422	16.6	mg/L	48	Standard
Fe	54	144.2	9.3	**-0.0052**	0.016	301.7	mg/L	159	Standard
Fe	57	226.7	13.5	**0.1451**	0.083	57.3	mg/L	195	Standard
Sc-1	45	39242.4	3.0				mg/L	39543	Standard
Cl	35	36295.4	2.3				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	2233.5	3.2				ug/L	1647	Standard
P	31	18441.2	3.0				ug/L	15231	Standard
S	34	2560.2	3.7				ug/L	1845	Standard
Sr	88	61.7	26.1				ug/L	70	Standard
C	12	130.0	20.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	22.9	112.2				mg/L	12	Standard
Ho-1	165	11.7	24.7				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	1601.8	24.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		89.555	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.037	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBS M4 WG574306-02
Report Date/Time: Tuesday, June 28, 2016 23:27:51
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.304
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.784
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBS M4 WG574306-02
Report Date/Time: Tuesday, June 28, 2016 23:27:51
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: LCSS M4 WG574306-03
Sample Date/Time: Tuesday, June 28, 2016 23:28:45
Number of Replicates: 3
Autosampler Position: 402
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72376.0	3.8				ug/L	81760	Standard
	Be	9	23423.1	1.4	**26.6532**	0.807	3.0	ug/L	17	Standard
	Al	27	9074.4	3.8	**0.0521**	0.003	4.9	ug/L	1410	Standard
	Sc	45	39197.3	2.4				ug/L	39543	Standard
	Ti	47	29.3	7.1	**-0.0442**	0.007	16.8	ug/L	39	Standard
	V	51	169885.7	0.8	**24.8594**	0.272	1.1	ug/L	1415	Standard
	Cr	52	181506.5	1.2	**25.6821**	0.081	0.3	ug/L	6946	Standard
	Cr	53	22394.9	2.0	**25.8190**	0.261	1.0	ug/L	328	Standard
	Mn	55	162938.5	1.1	**25.6253**	0.137	0.5	ug/L	823	Standard
	Co	59	171171.6	1.4	**24.9527**	0.363	1.5	ug/L	252	Standard
	Ni	60	47214.1	1.2	**25.6031**	0.067	0.3	ug/L	192	Standard
	Cu	65	47011.2	0.9	**25.2811**	0.042	0.2	ug/L	141	Standard
	Zn	66	28788.3	1.3	**26.1066**	0.165	0.6	ug/L	491	Standard
>	**Ge**	72	502290.9	1.0				ug/L	515087	Standard
	As	75	27387.3	0.6	**25.0208**	0.130	0.5	ug/L	20	Standard
	Se	82	2754.4	0.7	**24.4154**	0.147	0.6	ug/L	15	Standard
	Se-1	77	1916.1	0.9	**25.3637**	0.129	0.5	ug/L	48	Standard
>	**Ga**	71	55.0	18.2				mg/L	90	Standard
	Rb	85	31.7	18.2				ug/L	23	Standard
	Y	89	439399.0	0.9				ug/L	448510	Standard
>	**Rh**	103	31.7	24.1				ug/L	15	Standard
	Mo	98	98.7	7.7	**0.0186**	0.002	12.0	ug/L	11	Standard
	Ag	107	158446.9	0.8	**24.3265**	0.298	1.2	ug/L	134	Standard
	Cd	111	51639.0	0.8	**24.9106**	0.321	1.3	mg/L	7	Standard
	Cd	114	123238.5	0.9	**24.4562**	0.275	1.1	ug/L	16	Standard
>	**In**	115	592677.3	1.9				ug/L	611117	Standard
	Sn	118	653.3	6.2	**0.0119**	0.005	41.0	ug/L	518	Standard
	Sb	123	110361.6	0.7	**24.1094**	0.310	1.3	ug/L	127	Standard
	Ba	135	48745.3	0.9	**24.3246**	0.340	1.4	ug/L	45	Standard
	Ce	140	51.7	39.1				ug/L	298	Standard
>	**Tb**	159	905963.3	1.2				ug/L	977942	Standard
	Ho	165	11.7	65.5				ug/L	12	Standard
	Tl	203	202415.8	0.8	**25.4126**	0.129	0.5	ug/L	31	Standard
	Tl	205	180512.3	0.1	**26.1008**	0.236	0.9	ug/L	20	Standard
	Pb	206	132596.7	0.6	**25.8949**	0.268	1.0	ug/L	286	Standard
	Pb	207	115831.1	1.1	**25.0809**	0.349	1.4	ug/L	257	Standard
	Pb	208	485807.7	1.5	**27.1937**	0.548	2.0	ug/L	956	Standard
	U	238	174705.5	0.6	**25.7950**	0.133	0.5	ug/L	17	Standard
>	**Bi**	209	526543.8	0.9				ug/L	557080	Standard

Sample ID: LCSS M4 WG574306-03
Report Date/Time: Tuesday, June 28, 2016 23:31:02
Page 1

Approved: June 29, 2016

	Na	23	0.0		**0.0050**			mg/L	0	Standard
	Mg	24	28.3	53.9	**-0.0192**	0.020	104.0	mg/L	32	Standard
	K	39	18.3	78.7	**-0.0794**	0.048	60.3	mg/L	30	Standard
	Ca	43	26.7	21.7	**-1.2726**	0.799	62.8	mg/L	48	Standard
	Fe	54	143.6	13.2	**-0.0059**	0.017	289.5	mg/L	159	Standard
	Fe	57	241.7	12.5	**0.2002**	0.121	60.5	mg/L	195	Standard
>	**Sc-1**	45	39197.3	2.4				mg/L	39543	Standard
	Cl	35	35988.7	2.8				ug/L	27642	Standard
	Kr	83	3.3	17.3				ug/L	2	Standard
	Br	81	1690.1	11.1				ug/L	1647	Standard
	P	31	17910.6	1.0				ug/L	15231	Standard
	S	34	2548.5	6.5				ug/L	1845	Standard
	Sr	88	75.0	24.0				ug/L	70	Standard
	C	12	120.0	16.7				mg/L	113	Standard
	N	14	0.0					mg/L	0	Standard
	Hg	202	3.3	173.2				mg/L	0	Standard
	Dy	164	16.5	68.3				mg/L	12	Standard
	Ho-1	165	11.7	65.5				mg/L	12	Standard
	Er	166	3.3	173.2				mg/L	20	Standard
	I	127	1618.4	15.4				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		88.522	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		97.516	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: LCSS M4 WG574306-03
Report Date/Time: Tuesday, June 28, 2016 23:31:02
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.983
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.518
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSS M4 WG574306-03
Report Date/Time: Tuesday, June 28, 2016 23:31:02
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606129707 WG574306-01
Sample Date/Time: Tuesday, June 28, 2016 23:31:57
Number of Replicates: 3
Autosampler Position: 403
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73039.0	1.3				ug/L	81760	Standard
	Be	9	626.7	10.9	**0.6973**	0.086	12.3	ug/L	17	Standard
	Al	27	6016122.3	2.0	**43.6794**	0.992	2.3	ug/L	1410	Standard
	Sc	45	43592.6	1.8				ug/L	39543	Standard
	Ti	47	1645.1	1.5	**6.4955**	0.085	1.3	ug/L	39	Standard
	V	51	103833.7	2.4	**14.8338**	0.288	1.9	ug/L	1415	Standard
	Cr	52	59851.7	1.4	**7.6005**	0.153	2.0	ug/L	6946	Standard
	Cr	53	7268.4	3.1	**7.9966**	0.305	3.8	ug/L	328	Standard
	Mn	55	3268160.2	1.2	**507.4707**	9.760	1.9	ug/L	823	Standard
	Co	59	39533.2	0.9	**5.6343**	0.098	1.7	ug/L	252	Standard
	Ni	60	18874.4	2.7	**9.9863**	0.299	3.0	ug/L	192	Standard
	Cu	65	69246.2	1.1	**36.6021**	0.305	0.8	ug/L	141	Standard
	Zn	66	86637.7	0.8	**78.7605**	0.745	0.9	ug/L	491	Standard
>	**Ge**	72	511586.4	0.8				ug/L	515087	Standard
	As	75	4823.9	1.5	**4.3600**	0.029	0.7	ug/L	20	Standard
	Se	82	100.3	6.0	**0.7578**	0.047	6.2	ug/L	15	Standard
	Se-1	77	133.3	1.1	**1.0631**	0.006	0.6	ug/L	48	Standard
>	**Ga**	71	7216.7	0.8				mg/L	90	Standard
	Rb	85	77067.2	0.1				ug/L	23	Standard
	Y	89	607493.8	2.8				ug/L	448510	Standard
>	**Rh**	103	25.0	34.6				ug/L	15	Standard
	Mo	98	512.0	2.8	**0.1759**	0.005	2.9	ug/L	11	Standard
	Ag	107	750.4	6.7	**0.0988**	0.008	7.8	ug/L	134	Standard
	Cd	111	815.5	1.0	**0.3898**	0.005	1.3	mg/L	7	Standard
	Cd	114	2038.2	2.1	**0.3989**	0.010	2.4	ug/L	16	Standard
>	**In**	115	593721.9	0.4				ug/L	611117	Standard
	Sn	118	941.7	10.4	**0.0620**	0.016	26.6	ug/L	518	Standard
	Sb	123	1005.8	16.4	**0.2066**	0.037	17.7	ug/L	127	Standard
	Ba	135	393236.3	1.1	**196.0335**	2.433	1.2	ug/L	45	Standard
	Ce	140	857283.3	2.5				ug/L	298	Standard
>	**Tb**	159	944884.3	0.7				ug/L	977942	Standard
	Ho	165	15139.2	2.1				ug/L	12	Standard
	Tl	203	479.0	10.1	**0.0546**	0.006	10.6	ug/L	31	Standard
	Tl	205	396.7	12.4	**0.0544**	0.008	14.1	ug/L	20	Standard
	Pb	206	209606.6	0.8	**40.4302**	0.428	1.1	ug/L	286	Standard
	Pb	207	173972.5	0.8	**37.2011**	0.310	0.8	ug/L	257	Standard
	Pb	208	734775.6	0.6	**40.6204**	0.824	2.0	ug/L	956	Standard
	U	238	3055.0	2.6	**0.4423**	0.014	3.2	ug/L	17	Standard
>	**Bi**	209	533598.4	1.5				ug/L	557080	Standard

Sample ID: L1606129707 WG574306-01
Report Date/Time: Tuesday, June 28, 2016 23:34:13
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**19.2926**	33.407	173.2	mg/L	0	Standard
Mg	24	75.0	26.7	**0.0334**	0.024	70.8	mg/L	32	Standard
K	39	210.0	4.8	**0.5021**	0.023	4.7	mg/L	30	Standard
Ca	43	75.0	29.1	**4.9537**	2.848	57.5	mg/L	48	Standard
Fe	54	6552.0	2.8	**5.2441**	0.230	4.4	mg/L	159	Standard
Fe	57	1841.8	2.3	**5.0156**	0.091	1.8	mg/L	195	Standard
> **Sc-1**	45	43592.6	1.8				mg/L	39543	Standard
Cl	35	34528.0	1.5				ug/L	27642	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	3080.3	6.6				ug/L	1647	Standard
P	31	18262.7	4.8				ug/L	15231	Standard
S	34	2550.2	4.9				ug/L	1845	Standard
Sr	88	106.7	7.2				ug/L	70	Standard
C	12	370.0	20.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	40.0				mg/L	0	Standard
Dy	164	25551.2	3.2				mg/L	12	Standard
Ho-1	165	15139.2	2.1				mg/L	12	Standard
Er	166	14156.6	1.1				mg/L	20	Standard
I	127	14104.9	1.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		89.333	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.320	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129707 WG574306-01
Report Date/Time: Tuesday, June 28, 2016 23:34:13
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.154
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.785
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606129708S WG574306-04
Sample Date/Time: Tuesday, June 28, 2016 23:35:08
Number of Replicates: 3
Autosampler Position: 404
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77425.7	0.5				ug/L	81760	Standard
	Be	9	24978.9	2.1	**26.5514**	0.597	2.2	ug/L	17	Standard
	Al	27	20723622.4	0.1	**141.9584**	0.685	0.5	ug/L	1410	Standard
	Sc	45	47225.2	3.2				ug/L	39543	Standard
	Ti	47	4499.3	2.9	**18.2724**	0.558	3.1	ug/L	39	Standard
	V	51	344846.0	0.3	**50.3538**	0.699	1.4	ug/L	1415	Standard
	Cr	52	312957.5	0.8	**44.7646**	1.096	2.4	ug/L	6946	Standard
	Cr	53	39581.6	1.4	**45.6150**	1.229	2.7	ug/L	328	Standard
	Mn	55	5573425.5	1.4	**875.8629**	6.350	0.7	ug/L	823	Standard
	Co	59	254404.2	1.0	**36.8653**	0.333	0.9	ug/L	252	Standard
	Ni	60	98601.5	0.1	**53.2500**	0.840	1.6	ug/L	192	Standard
	Cu	65	83052.3	0.8	**44.4536**	0.827	1.9	ug/L	141	Standard
	Zn	66	91760.7	0.8	**84.4843**	0.972	1.2	ug/L	491	Standard
>	**Ge**	72	505521.9	1.6				ug/L	515087	Standard
	As	75	29384.3	0.4	**26.6737**	0.341	1.3	ug/L	20	Standard
	Se	82	1902.3	1.0	**16.7176**	0.177	1.1	ug/L	15	Standard
	Se-1	77	1368.1	4.0	**17.7813**	0.553	3.1	ug/L	48	Standard
>	**Ga**	71	22081.1	1.8				mg/L	90	Standard
	Rb	85	217862.7	3.0				ug/L	23	Standard
	Y	89	681158.2	0.5				ug/L	448510	Standard
>	**Rh**	103	41.7	13.9				ug/L	15	Standard
	Mo	98	502.0	0.9	**0.1763**	0.003	1.6	ug/L	11	Standard
	Ag	107	151390.6	0.2	**23.7106**	0.326	1.4	ug/L	134	Standard
	Cd	111	50168.4	0.1	**24.6878**	0.285	1.2	mg/L	7	Standard
	Cd	114	122073.8	0.4	**24.7133**	0.372	1.5	ug/L	16	Standard
>	**In**	115	580961.0	1.3				ug/L	611117	Standard
	Sn	118	790.0	7.7	**0.0385**	0.009	23.5	ug/L	518	Standard
	Sb	123	3165.4	5.9	**0.6928**	0.037	5.4	ug/L	127	Standard
	Ba	135	532338.7	0.4	**271.2464**	4.110	1.5	ug/L	45	Standard
	Ce	140	1309534.5	1.1				ug/L	298	Standard
>	**Tb**	159	945803.3	0.3				ug/L	977942	Standard
	Ho	165	21665.5	2.8				ug/L	12	Standard
	Tl	203	195988.5	0.7	**24.3661**	0.181	0.7	ug/L	31	Standard
	Tl	205	172319.6	2.0	**24.6721**	0.452	1.8	ug/L	20	Standard
	Pb	206	254741.0	0.7	**49.3204**	0.423	0.9	ug/L	286	Standard
	Pb	207	216152.1	1.4	**46.3962**	0.722	1.6	ug/L	257	Standard
	Pb	208	888930.7	0.3	**49.3200**	0.099	0.2	ug/L	956	Standard
	U	238	171966.4	0.9	**25.1430**	0.188	0.7	ug/L	17	Standard
>	**Bi**	209	531711.1	0.2				ug/L	557080	Standard

Sample ID: L1606129708S WG574306-04
Report Date/Time: Tuesday, June 28, 2016 23:37:25
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	125.0	34.9	**0.0812**	0.045	55.3	mg/L	32	Standard
K	39	518.3	8.6	**1.3288**	0.177	13.3	mg/L	30	Standard
Ca	43	238.3	9.5	**24.9959**	3.131	12.5	mg/L	48	Standard
Fe	54	18930.5	0.8	**14.2167**	0.407	2.9	mg/L	159	Standard
Fe	57	5154.2	3.0	**13.9467**	0.213	1.5	mg/L	195	Standard
> **Sc-1**	45	47225.2	3.2				mg/L	39543	Standard
Cl	35	33388.1	1.2				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	2326.8	4.2				ug/L	1647	Standard
P	31	19087.0	1.5				ug/L	15231	Standard
S	34	2648.6	1.1				ug/L	1845	Standard
Sr	88	121.7	17.1				ug/L	70	Standard
C	12	206.7	36.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	36.7	56.8				mg/L	0	Standard
Dy	164	35480.3	0.5				mg/L	12	Standard
Ho-1	165	21665.5	2.8				mg/L	12	Standard
Er	166	19556.0	2.5				mg/L	20	Standard
I	127	13694.5	2.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.699	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.143	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129708S WG574306-04
Report Date/Time: Tuesday, June 28, 2016 23:37:25
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	95.065
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.446
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606129708S WG574306-04
Report Date/Time: Tuesday, June 28, 2016 23:37:25
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606129709SD WG574306-05
Sample Date/Time: Tuesday, June 28, 2016 23:38:19
Number of Replicates: 3
Autosampler Position: 405
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74161.1	1.1				ug/L	81760	Standard
	Be	9	23705.2	3.8	**26.3039**	0.872	3.3	ug/L	17	Standard
	Al	27	6111125.0	1.3	**43.6951**	0.516	1.2	ug/L	1410	Standard
	Sc	45	44620.6	1.2				ug/L	39543	Standard
	Ti	47	2932.6	3.4	**11.8635**	0.492	4.1	ug/L	39	Standard
	V	51	246865.2	0.2	**36.0158**	0.450	1.2	ug/L	1415	Standard
	Cr	52	213059.8	1.6	**30.1549**	0.173	0.6	ug/L	6946	Standard
	Cr	53	26094.1	1.0	**29.9771**	0.650	2.2	ug/L	328	Standard
	Mn	55	1521366.7	1.6	**239.1613**	2.815	1.2	ug/L	823	Standard
	Co	59	194513.2	1.8	**28.2005**	0.410	1.5	ug/L	252	Standard
	Ni	60	61728.2	1.1	**33.3195**	0.163	0.5	ug/L	192	Standard
	Cu	65	74436.8	1.3	**39.8619**	0.705	1.8	ug/L	141	Standard
	Zn	66	58904.3	0.8	**53.9789**	0.545	1.0	ug/L	491	Standard
>	**Ge**	72	505111.3	1.2				ug/L	515087	Standard
	As	75	31837.0	1.5	**28.9187**	0.503	1.7	ug/L	20	Standard
	Se	82	2328.1	1.1	**20.5054**	0.470	2.3	ug/L	15	Standard
	Se-1	77	1619.4	1.4	**21.2042**	0.440	2.1	ug/L	48	Standard
>	**Ga**	71	6334.7	3.0				mg/L	90	Standard
	Rb	85	74980.1	1.6				ug/L	23	Standard
	Y	89	605551.0	0.9				ug/L	448510	Standard
>	**Rh**	103	31.7	39.7				ug/L	15	Standard
	Mo	98	404.4	5.7	**0.1335**	0.008	5.8	ug/L	11	Standard
	Ag	107	149725.6	1.6	**22.7377**	0.610	2.7	ug/L	134	Standard
	Cd	111	49979.3	1.2	**23.8462**	0.386	1.6	mg/L	7	Standard
	Cd	114	121386.1	1.5	**23.8269**	0.554	2.3	ug/L	16	Standard
>	**In**	115	599198.2	1.3				ug/L	611117	Standard
	Sn	118	811.7	7.5	**0.0382**	0.012	32.5	ug/L	518	Standard
	Sb	123	19138.7	1.5	**4.1252**	0.117	2.8	ug/L	127	Standard
	Ba	135	258930.8	0.8	**127.9040**	1.896	1.5	ug/L	45	Standard
	Ce	140	760518.7	0.8				ug/L	298	Standard
>	**Tb**	159	951283.8	0.9				ug/L	977942	Standard
	Ho	165	14106.6	0.1				ug/L	12	Standard
	Tl	203	188730.2	0.5	**23.1187**	0.087	0.4	ug/L	31	Standard
	Tl	205	170134.9	1.0	**24.0027**	0.337	1.4	ug/L	20	Standard
	Pb	206	208981.3	1.7	**39.8515**	0.361	0.9	ug/L	286	Standard
	Pb	207	178505.4	0.4	**37.7421**	0.330	0.9	ug/L	257	Standard
	Pb	208	755416.1	1.2	**41.2868**	0.479	1.2	ug/L	956	Standard
	U	238	170025.1	1.0	**24.4935**	0.095	0.4	ug/L	17	Standard
>	**Bi**	209	539649.3	0.9				ug/L	557080	Standard

Sample ID: L1606129709SD WG574306-05
Report Date/Time: Tuesday, June 28, 2016 23:40:36
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**19.0051**	16.458	86.6	mg/L	0	Standard
Mg	24	83.3	27.7	**0.0410**	0.026	64.1	mg/L	32	Standard
K	39	181.7	7.9	**0.4027**	0.045	11.2	mg/L	30	Standard
Ca	43	75.0	26.7	**4.7351**	2.622	55.4	mg/L	48	Standard
Fe	54	5650.5	3.5	**4.3963**	0.211	4.8	mg/L	159	Standard
Fe	57	1623.4	7.9	**4.2316**	0.365	8.6	mg/L	195	Standard
> **Sc-1**	45	44620.6	1.2				mg/L	39543	Standard
Cl	35	32136.0	1.3				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1900.1	7.1				ug/L	1647	Standard
P	31	17840.5	2.2				ug/L	15231	Standard
S	34	2363.5	5.1				ug/L	1845	Standard
Sr	88	98.3	20.5				ug/L	70	Standard
C	12	150.0	37.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	20.0				mg/L	0	Standard
Dy	164	23948.7	0.5				mg/L	12	Standard
Ho-1	165	14106.6	0.1				mg/L	12	Standard
Er	166	12898.8	4.1				mg/L	20	Standard
I	127	5567.7	3.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		90.706	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.063	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129709SD WG574306-05
Report Date/Time: Tuesday, June 28, 2016 23:40:36
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.050
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	96.871
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606129709SD WG574306-05
Report Date/Time: Tuesday, June 28, 2016 23:40:36
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606144301
Sample Date/Time: Tuesday, June 28, 2016 23:41:31
Number of Replicates: 3
Autosampler Position: 406
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72988.8	2.0				ug/L	81760	Standard
	Be	9	916.7	4.9	**1.0251**	0.070	6.9	ug/L	17	Standard
	Al	27	29327581.2	1.7	**213.1994**	7.066	3.3	ug/L	1410	Standard
	Sc	45	42180.3	1.1				ug/L	39543	Standard
	Ti	47	15217.0	2.1	**62.4530**	1.008	1.6	ug/L	39	Standard
	V	51	58550.8	0.3	**8.4144**	0.093	1.1	ug/L	1415	Standard
	Cr	52	50375.3	2.1	**6.3484**	0.119	1.9	ug/L	6946	Standard
	Cr	53	5892.8	2.4	**6.5282**	0.148	2.3	ug/L	328	Standard
	Mn	55	1501608.8	1.0	**236.8837**	1.387	0.6	ug/L	823	Standard
	Co	59	11082.4	1.1	**1.5829**	0.023	1.4	ug/L	252	Standard
	Ni	60	9938.9	1.0	**5.2962**	0.105	2.0	ug/L	192	Standard
	Cu	65	12880.8	1.6	**6.8461**	0.032	0.5	ug/L	141	Standard
	Zn	66	52551.7	1.1	**48.2375**	0.337	0.7	ug/L	491	Standard
>	**Ge**	72	503346.2	1.1				ug/L	515087	Standard
	As	75	9638.5	1.8	**8.8128**	0.107	1.2	ug/L	20	Standard
	Se	82	89.1	11.5	**0.6728**	0.088	13.1	ug/L	15	Standard
	Se-1	77	112.0	20.6	**0.8047**	0.328	40.8	ug/L	48	Standard
>	**Ga**	71	5402.6	1.6				mg/L	90	Standard
	Rb	85	27092.5	2.9				ug/L	23	Standard
	Y	89	512619.4	0.7				ug/L	448510	Standard
>	**Rh**	103	40.0	45.1				ug/L	15	Standard
	Mo	98	2129.4	2.5	**0.7922**	0.032	4.0	ug/L	11	Standard
	Ag	107	397.3	28.5	**0.0446**	0.016	36.8	ug/L	134	Standard
	Cd	111	847.6	0.8	**0.4054**	0.009	2.1	mg/L	7	Standard
	Cd	114	2047.7	3.4	**0.4008**	0.011	2.7	ug/L	16	Standard
>	**In**	115	593663.8	1.5				ug/L	611117	Standard
	Sn	118	1955.1	5.4	**0.2386**	0.021	8.9	ug/L	518	Standard
	Sb	123	857.9	5.2	**0.1744**	0.012	7.0	ug/L	127	Standard
	Ba	135	91007.2	1.4	**45.3648**	1.263	2.8	ug/L	45	Standard
	Ce	140	427372.5	1.2				ug/L	298	Standard
>	**Tb**	159	914890.2	3.3				ug/L	977942	Standard
	Ho	165	5369.3	2.3				ug/L	12	Standard
	Tl	203	3260.4	2.3	**0.4133**	0.015	3.6	ug/L	31	Standard
	Tl	205	2856.9	4.2	**0.4193**	0.011	2.7	ug/L	20	Standard
	Pb	206	241551.6	0.7	**48.2276**	1.641	3.4	ug/L	286	Standard
	Pb	207	199984.5	0.7	**44.2621**	1.381	3.1	ug/L	257	Standard
	Pb	208	845904.8	0.5	**48.3946**	1.395	2.9	ug/L	956	Standard
	U	238	4244.6	0.7	**0.6370**	0.016	2.5	ug/L	17	Standard
>	**Bi**	209	515907.9	2.8				ug/L	557080	Standard

Sample ID: L1606144301
Report Date/Time: Tuesday, June 28, 2016 23:43:48
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.1607**	17.590	173.1	mg/L	0	Standard
Mg	24	176.7	13.4	**0.1635**	0.031	19.2	mg/L	32	Standard
K	39	111.7	29.8	**0.2126**	0.107	50.1	mg/L	30	Standard
Ca	43	486.7	20.1	**63.9109**	13.520	21.2	mg/L	48	Standard
Fe	54	4570.1	3.9	**3.7396**	0.111	3.0	mg/L	159	Standard
Fe	57	1611.8	3.6	**4.4774**	0.217	4.9	mg/L	195	Standard
> **Sc-1**	45	42180.3	1.1				mg/L	39543	Standard
Cl	35	34952.3	0.5				ug/L	27642	Standard
Kr	83	2.3	65.5				ug/L	2	Standard
Br	81	2820.3	6.2				ug/L	1647	Standard
P	31	18733.3	7.5				ug/L	15231	Standard
S	34	2421.9	7.6				ug/L	1845	Standard
Sr	88	105.0	16.5				ug/L	70	Standard
C	12	266.7	9.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	83.3	60.4				mg/L	0	Standard
Dy	164	8334.8	4.4				mg/L	12	Standard
Ho-1	165	5369.3	2.3				mg/L	12	Standard
Er	166	5100.9	4.8				mg/L	20	Standard
I	127	10480.3	8.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		89.272	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.721	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144301
Report Date/Time: Tuesday, June 28, 2016 23:43:48
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.144
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	92.609
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606144302
Sample Date/Time: Tuesday, June 28, 2016 23:44:42
Number of Replicates: 3
Autosampler Position: 407
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	71243.8	1.4				ug/L	81760	Standard
	Be	9	1080.0	8.5	**1.2377**	0.094	7.6	ug/L	17	Standard
	Al	27	5504266.2	2.7	**40.9647**	0.928	2.3	ug/L	1410	Standard
	Sc	45	42188.7	3.4				ug/L	39543	Standard
	Ti	47	15737.8	0.6	**62.0152**	0.652	1.1	ug/L	39	Standard
	V	51	45512.4	1.5	**6.2276**	0.158	2.5	ug/L	1415	Standard
	Cr	52	44821.9	1.8	**5.2702**	0.176	3.3	ug/L	6946	Standard
	Cr	53	5345.9	2.2	**5.6430**	0.187	3.3	ug/L	328	Standard
	Mn	55	1510551.7	1.4	**228.8078**	5.482	2.4	ug/L	823	Standard
	Co	59	8314.6	0.3	**1.1314**	0.011	1.0	ug/L	252	Standard
	Ni	60	7372.1	1.0	**3.7410**	0.046	1.2	ug/L	192	Standard
	Cu	65	13129.7	1.4	**6.6988**	0.146	2.2	ug/L	141	Standard
	Zn	66	50712.1	0.8	**44.6331**	0.723	1.6	ug/L	491	Standard
>	**Ge**	72	524277.6	1.0				ug/L	515087	Standard
	As	75	7848.1	0.4	**6.8989**	0.093	1.3	ug/L	20	Standard
	Se	82	94.2	13.2	**0.6849**	0.109	15.9	ug/L	15	Standard
	Se-1	77	110.0	9.5	**0.7149**	0.123	17.2	ug/L	48	Standard
>	**Ga**	71	4740.7	7.0				mg/L	90	Standard
	Rb	85	22735.4	3.5				ug/L	23	Standard
	Y	89	510942.2	1.8				ug/L	448510	Standard
>	**Rh**	103	18.3	110.2				ug/L	15	Standard
	Mo	98	3616.2	2.9	**1.3869**	0.053	3.8	ug/L	11	Standard
	Ag	107	415.3	27.2	**0.0487**	0.017	35.1	ug/L	134	Standard
	Cd	111	879.6	1.3	**0.4294**	0.003	0.8	mg/L	7	Standard
	Cd	114	2156.4	4.6	**0.4311**	0.017	3.9	ug/L	16	Standard
>	**In**	115	581659.3	1.0				ug/L	611117	Standard
	Sn	118	1941.8	8.1	**0.2429**	0.025	10.2	ug/L	518	Standard
	Sb	123	795.2	3.5	**0.1642**	0.005	3.1	ug/L	127	Standard
	Ba	135	105670.2	0.5	**53.7566**	0.798	1.5	ug/L	45	Standard
	Ce	140	489350.3	0.7				ug/L	298	Standard
>	**Tb**	159	914906.3	0.9				ug/L	977942	Standard
	Ho	165	5852.8	3.2				ug/L	12	Standard
	Tl	203	2193.8	7.0	**0.2695**	0.018	6.5	ug/L	31	Standard
	Tl	205	1858.4	8.2	**0.2655**	0.024	8.9	ug/L	20	Standard
	Pb	206	250441.6	1.0	**48.7728**	1.107	2.3	ug/L	286	Standard
	Pb	207	208820.1	0.2	**45.0828**	0.807	1.8	ug/L	257	Standard
	Pb	208	871187.9	1.2	**48.6248**	1.437	3.0	ug/L	956	Standard
	U	238	4748.1	0.9	**0.6954**	0.016	2.3	ug/L	17	Standard
>	**Bi**	209	528719.9	1.7				ug/L	557080	Standard

Sample ID: L1606144302
Report Date/Time: Tuesday, June 28, 2016 23:46:59
Page 1

Approved: June 29, 2016

Na	23	6.7	43.3	**39.8641**	16.152	40.5	mg/L	0	Standard
Mg	24	568.3	2.7	**0.6523**	0.023	3.6	mg/L	32	Standard
K	39	88.3	11.8	**0.1385**	0.034	24.4	mg/L	30	Standard
Ca	43	236.7	18.2	**28.4335**	6.608	23.2	mg/L	48	Standard
Fe	54	3607.3	1.4	**2.9257**	0.132	4.5	mg/L	159	Standard
Fe	57	1208.4	3.1	**3.2044**	0.245	7.6	mg/L	195	Standard
> **Sc-1**	45	42188.7	3.4				mg/L	39543	Standard
Cl	35	33795.0	1.6				ug/L	27642	Standard
Kr	83	3.3	75.5				ug/L	2	Standard
Br	81	2200.2	4.1				ug/L	1647	Standard
P	31	17892.2	3.9				ug/L	15231	Standard
S	34	2398.5	0.8				ug/L	1845	Standard
Sr	88	86.7	23.3				ug/L	70	Standard
C	12	246.7	15.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	76.7	30.1				mg/L	0	Standard
Dy	164	8934.1	5.5				mg/L	12	Standard
Ho-1	165	5852.8	3.2				mg/L	12	Standard
Er	166	5817.8	2.9				mg/L	20	Standard
I	127	10927.3	3.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		87.138	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.784	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144302
Report Date/Time: Tuesday, June 28, 2016 23:46:59
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.180
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.909
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606144302PS WG574573-01
Sample Date/Time: Tuesday, June 28, 2016 23:47:54
Number of Replicates: 3
Autosampler Position: 408
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	74517.8	1.0				ug/L	81760	Standard
	Be	9	49025.9	0.9	**54.1590**	0.839	1.5	ug/L	17	Standard
	Al	27	5488864.5	1.0	**39.0606**	0.777	2.0	ug/L	1410	Standard
	Sc	45	43992.1	1.6				ug/L	39543	Standard
	Ti	47	15590.7	1.3	**62.0794**	2.775	4.5	ug/L	39	Standard
	V	51	384779.8	0.6	**54.7401**	1.764	3.2	ug/L	1415	Standard
	Cr	52	400124.0	1.3	**56.0091**	2.445	4.4	ug/L	6946	Standard
	Cr	53	50664.6	1.6	**56.9262**	0.956	1.7	ug/L	328	Standard
	Mn	55	1785071.4	0.4	**273.1709**	9.729	3.6	ug/L	823	Standard
	Co	59	343226.4	0.5	**48.4551**	1.560	3.2	ug/L	252	Standard
	Ni	60	101539.8	1.7	**53.4214**	2.390	4.5	ug/L	192	Standard
	Cu	65	107630.3	0.5	**56.1285**	1.666	3.0	ug/L	141	Standard
	Zn	66	105446.7	0.4	**94.6554**	2.823	3.0	ug/L	491	Standard
>	Ge	72	519316.9	3.3				ug/L	515087	Standard
	As	75	63635.8	1.4	**56.2203**	2.024	3.6	ug/L	20	Standard
	Se	82	5672.0	2.3	**48.8026**	2.645	5.4	ug/L	15	Standard
	Se-1	77	3855.8	0.1	**50.0819**	1.642	3.3	ug/L	48	Standard
>	Ga	71	4749.1	6.7				mg/L	90	Standard
	Rb	85	22398.2	1.3				ug/L	23	Standard
	Y	89	511504.7	2.1				ug/L	448510	Standard
>	Rh	103	58.3	34.6				ug/L	15	Standard
	Mo	98	3550.9	0.5	**1.3279**	0.015	1.1	ug/L	11	Standard
	Ag	107	317741.2	0.8	**48.5166**	0.462	1.0	ug/L	134	Standard
	Cd	111	103027.6	1.2	**49.4156**	0.686	1.4	mg/L	7	Standard
	Cd	114	250223.1	1.7	**49.3820**	1.399	2.8	ug/L	16	Standard
>	In	115	596094.6	1.6				ug/L	611117	Standard
	Sn	118	2016.8	5.9	**0.2476**	0.017	6.7	ug/L	518	Standard
	Sb	123	226110.3	0.9	**49.1269**	0.970	2.0	ug/L	127	Standard
	Ba	135	199376.6	0.4	**98.9961**	1.356	1.4	ug/L	45	Standard
	Ce	140	473631.7	1.6				ug/L	298	Standard
>	Tb	159	946510.9	0.6				ug/L	977942	Standard
	Ho	165	5706.1	1.8				ug/L	12	Standard
	Tl	203	404565.2	1.6	**49.2192**	0.432	0.9	ug/L	31	Standard
	Tl	205	358549.4	2.4	**50.2279**	0.323	0.6	ug/L	20	Standard
	Pb	206	506766.3	0.6	**96.0764**	1.688	1.8	ug/L	286	Standard
	Pb	207	430425.5	0.4	**90.4686**	1.648	1.8	ug/L	257	Standard
	Pb	208	1789019.3	1.0	**97.1851**	0.775	0.8	ug/L	956	Standard
	U	238	348425.0	0.3	**49.8561**	0.765	1.5	ug/L	17	Standard
>	Bi	209	543434.9	1.8				ug/L	557080	Standard

Sample ID: L1606144302PS WG574573-01
Report Date/Time: Tuesday, June 28, 2016 23:50:11
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**19.2423**	16.666	86.6	mg/L	0	Standard
Mg	24	530.0	3.8	**0.5773**	0.030	5.2	mg/L	32	Standard
K	39	95.0	5.3	**0.1471**	0.014	9.3	mg/L	30	Standard
Ca	43	235.0	26.6	**26.8058**	8.969	33.5	mg/L	48	Standard
Fe	54	3632.6	2.6	**2.8189**	0.112	4.0	mg/L	159	Standard
Fe	57	1290.1	7.8	**3.2920**	0.337	10.2	mg/L	195	Standard
> **Sc-1**	45	43992.1	1.6				mg/L	39543	Standard
Cl	35	33127.5	1.3				ug/L	27642	Standard
Kr	83	2.0	100.0				ug/L	2	Standard
Br	81	2050.1	2.6				ug/L	1647	Standard
P	31	18372.8	5.1				ug/L	15231	Standard
S	34	2410.2	0.9				ug/L	1845	Standard
Sr	88	108.3	17.5				ug/L	70	Standard
C	12	236.7	47.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	86.7	70.5				mg/L	0	Standard
Dy	164	8282.4	2.8				mg/L	12	Standard
Ho-1	165	5706.1	1.8				mg/L	12	Standard
Er	166	5427.6	1.3				mg/L	20	Standard
I	127	10600.4	5.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.142	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.821	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144302PS WG574573-01
Report Date/Time: Tuesday, June 28, 2016 23:50:11
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.542
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.551
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606144302PS WG574573-01
Report Date/Time: Tuesday, June 28, 2016 23:50:11
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606144302SDL WG574573-02
Sample Date/Time: Tuesday, June 28, 2016 23:51:05
Number of Replicates: 3
Autosampler Position: 409
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	73539.7	0.5				ug/L	81760	Standard
	Be	9	228.3	17.7	**0.2458**	0.045	18.1	ug/L	17	Standard
	Al	27	1128305.0	2.6	**8.1231**	0.173	2.1	ug/L	1410	Standard
	Sc	45	40995.4	3.1				ug/L	39543	Standard
	Ti	47	3133.3	3.6	**12.6953**	0.331	2.6	ug/L	39	Standard
	V	51	9700.8	3.1	**1.2183**	0.050	4.1	ug/L	1415	Standard
	Cr	52	13743.6	3.1	**0.9582**	0.044	4.6	ug/L	6946	Standard
	Cr	53	1285.1	5.5	**1.1537**	0.099	8.6	ug/L	328	Standard
	Mn	55	302649.7	2.2	**47.5055**	0.538	1.1	ug/L	823	Standard
	Co	59	1856.8	5.3	**0.2384**	0.012	4.9	ug/L	252	Standard
	Ni	60	1619.8	3.4	**0.7729**	0.024	3.1	ug/L	192	Standard
	Cu	65	2766.6	1.1	**1.3956**	0.008	0.6	ug/L	141	Standard
	Zn	66	11859.0	1.8	**10.2169**	0.056	0.5	ug/L	491	Standard
>	Ge	72	504593.5	1.7				ug/L	515087	Standard
	As	75	1530.8	2.0	**1.4307**	0.048	3.3	ug/L	20	Standard
	Se	82	33.4	31.2	**0.1774**	0.094	53.0	ug/L	15	Standard
	Se-1	77	64.7	12.6	**0.1577**	0.113	71.7	ug/L	48	Standard
>	Ga	71	975.0	10.5				mg/L	90	Standard
	Rb	85	4577.4	4.1				ug/L	23	Standard
	Y	89	452862.0	1.2				ug/L	448510	Standard
>	Rh	103	23.3	32.7				ug/L	15	Standard
	Mo	98	722.3	4.5	**0.2491**	0.011	4.3	ug/L	11	Standard
	Ag	107	159.0	11.1	**0.0076**	0.003	35.4	ug/L	134	Standard
	Cd	111	175.3	14.8	**0.0794**	0.012	14.8	mg/L	7	Standard
	Cd	114	462.3	4.9	**0.0844**	0.004	4.5	ug/L	16	Standard
>	In	115	608749.5	0.7				ug/L	611117	Standard
	Sn	118	845.0	9.9	**0.0416**	0.014	34.5	ug/L	518	Standard
	Sb	123	2112.7	19.6	**0.4370**	0.091	20.8	ug/L	127	Standard
	Ba	135	21046.6	3.7	**10.2079**	0.317	3.1	ug/L	45	Standard
	Ce	140	96335.3	1.6				ug/L	298	Standard
>	Tb	159	932545.8	1.3				ug/L	977942	Standard
	Ho	165	1166.7	10.3				ug/L	12	Standard
	Tl	203	511.3	6.9	**0.0588**	0.004	6.8	ug/L	31	Standard
	Tl	205	456.7	14.9	**0.0631**	0.009	14.8	ug/L	20	Standard
	Pb	206	50001.0	0.8	**9.6268**	0.026	0.3	ug/L	286	Standard
	Pb	207	41687.9	1.1	**8.8985**	0.042	0.5	ug/L	257	Standard
	Pb	208	175842.9	0.5	**9.7051**	0.030	0.3	ug/L	956	Standard
	U	238	977.7	3.5	**0.1400**	0.006	4.1	ug/L	17	Standard
>	Bi	209	531859.1	0.7				ug/L	557080	Standard

Sample ID: L1606144302SDL WG574573-02
Report Date/Time: Tuesday, June 28, 2016 23:53:22
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**20.9199**	18.126	86.6	mg/L	0	Standard
Mg	24	148.3	20.6	**0.1326**	0.034	25.5	mg/L	32	Standard
K	39	35.0	24.7	**-0.0267**	0.032	120.4	mg/L	30	Standard
Ca	43	75.0	20.0	**5.6602**	2.270	40.1	mg/L	48	Standard
Fe	54	775.4	6.3	**0.5410**	0.062	11.5	mg/L	159	Standard
Fe	57	435.0	3.0	**0.7919**	0.003	0.4	mg/L	195	Standard
> **Sc-1**	45	40995.4	3.1				mg/L	39543	Standard
Cl	35	32746.0	2.5				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1743.4	12.5				ug/L	1647	Standard
P	31	17056.3	1.5				ug/L	15231	Standard
S	34	2336.8	6.2				ug/L	1845	Standard
Sr	88	78.3	20.5				ug/L	70	Standard
C	12	126.7	27.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	89.2				mg/L	0	Standard
Dy	164	1870.0	3.5				mg/L	12	Standard
Ho-1	165	1166.7	10.3				mg/L	12	Standard
Er	166	1116.7	18.2				mg/L	20	Standard
I	127	3388.7	11.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		89.946	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.963	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144302SDL WG574573-02
Report Date/Time: Tuesday, June 28, 2016 23:53:22
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.613
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.473
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606144302SDL WG574573-02
Sample Date/Time: Tuesday, June 28, 2016 23:54:17
Number of Replicates: 3
Autosampler Position: 410
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70935.7	2.2				ug/L	81760	Standard
	Be	9	55.0	15.7	**0.0540**	0.009	16.0	ug/L	17	Standard
	Al	27	218570.5	1.6	**1.6201**	0.011	0.7	ug/L	1410	Standard
	Sc	45	39984.4	1.2				ug/L	39543	Standard
	Ti	47	627.7	4.3	**2.4198**	0.114	4.7	ug/L	39	Standard
	V	51	2847.0	3.0	**0.2129**	0.013	6.1	ug/L	1415	Standard
	Cr	52	7618.6	1.2	**0.0637**	0.011	17.1	ug/L	6946	Standard
	Cr	53	580.0	9.6	**0.3354**	0.068	20.2	ug/L	328	Standard
	Mn	55	61547.4	1.3	**9.5745**	0.188	2.0	ug/L	823	Standard
	Co	59	523.0	3.3	**0.0449**	0.003	6.0	ug/L	252	Standard
	Ni	60	493.7	5.6	**0.1635**	0.016	10.1	ug/L	192	Standard
	Cu	65	675.0	3.2	**0.2729**	0.010	3.6	ug/L	141	Standard
	Zn	66	4565.4	1.3	**3.4380**	0.075	2.2	ug/L	491	Standard
>	**Ge**	72	502988.1	0.7				ug/L	515087	Standard
	As	75	269.2	10.7	**0.2856**	0.028	9.7	ug/L	20	Standard
	Se	82	20.7	22.3	**0.0646**	0.042	64.7	ug/L	15	Standard
	Se-1	77	45.7	15.4	**-0.0984**	0.093	94.5	ug/L	48	Standard
>	**Ga**	71	186.7	12.1				mg/L	90	Standard
	Rb	85	835.0	12.4				ug/L	23	Standard
	Y	89	441201.2	2.7				ug/L	448510	Standard
>	**Rh**	103	13.3	94.4				ug/L	15	Standard
	Mo	98	189.0	7.0	**0.0532**	0.006	12.0	ug/L	11	Standard
	Ag	107	91.3	11.0	**-0.0021**	0.002	87.0	ug/L	134	Standard
	Cd	111	37.0	12.7	**0.0150**	0.002	15.9	mg/L	7	Standard
	Cd	114	85.9	30.8	**0.0122**	0.005	43.3	ug/L	16	Standard
>	**In**	115	591854.8	2.3				ug/L	611117	Standard
	Sn	118	648.3	6.3	**0.0113**	0.007	65.4	ug/L	518	Standard
	Sb	123	522.8	33.2	**0.1022**	0.041	39.7	ug/L	127	Standard
	Ba	135	4107.2	1.3	**2.0294**	0.025	1.2	ug/L	45	Standard
	Ce	140	18988.6	1.0				ug/L	298	Standard
>	**Tb**	159	914076.3	1.3				ug/L	977942	Standard
	Ho	165	253.3	13.1				ug/L	12	Standard
	Tl	203	112.3	10.6	**0.0094**	0.002	17.3	ug/L	31	Standard
	Tl	205	90.0	29.4	**0.0108**	0.004	36.1	ug/L	20	Standard
	Pb	206	10230.5	2.1	**1.9438**	0.049	2.5	ug/L	286	Standard
	Pb	207	8531.4	1.5	**1.7976**	0.010	0.6	ug/L	257	Standard
	Pb	208	36086.5	0.6	**1.9686**	0.029	1.5	ug/L	956	Standard
	U	238	209.0	10.0	**0.0280**	0.003	11.9	ug/L	17	Standard
>	**Bi**	209	525348.3	0.9				ug/L	557080	Standard

Sample ID: L1606144302SDL WG574573-02
Report Date/Time: Tuesday, June 28, 2016 23:56:34
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	68.3	18.4	**0.0329**	0.017	52.2	mg/L	32	Standard
K	39	41.7	42.1	**-0.0025**	0.057	2323.2	mg/L	30	Standard
Ca	43	35.0	49.5	**-0.0873**	2.606	2984.1	mg/L	48	Standard
Fe	54	276.1	10.3	**0.1102**	0.028	25.1	mg/L	159	Standard
Fe	57	273.3	14.9	**0.2889**	0.146	50.5	mg/L	195	Standard
> **Sc-1**	45	39984.4	1.2				mg/L	39543	Standard
Cl	35	32328.5	1.4				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	1603.4	8.8				ug/L	1647	Standard
P	31	16148.6	1.7				ug/L	15231	Standard
S	34	2251.8	6.8				ug/L	1845	Standard
Sr	88	76.7	35.9				ug/L	70	Standard
C	12	146.7	10.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	330.4	18.3				mg/L	12	Standard
Ho-1	165	253.3	13.1				mg/L	12	Standard
Er	166	270.0	22.5				mg/L	20	Standard
I	127	1720.1	18.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		86.761	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.651	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144302SDL WG574573-02
Report Date/Time: Tuesday, June 28, 2016 23:56:34
Page 2

Approved: June 29, 2016

⌊	Rb	85	
⌈	Y	89	
⌊>	Rh	103	
⌈	Mo	98	
|	Ag	107	
|	Cd	111	
|	Cd	114	
|>	In	115	96.848
|	Sn	118	
|	Sb	123	
⌊	Ba	135	
⌈	Ce	140	
⌊>	Tb	159	
⌈	Ho	165	
|	Tl	203	
|	Tl	205	
|	Pb	206	
|	Pb	207	
|	Pb	208	
|	U	238	
⌊>	Bi	209	94.304
⌈	Na	23	
|	Mg	24	
|	K	39	
|	Ca	43	
|	Fe	54	
|	Fe	57	
⌊>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 23:57:30
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	72062.8	4.4				ug/L	81760	Standard
	Be	9	45152.2	3.9	**51.5939**	1.461	2.8	ug/L	17	Standard
	Al	27	7383002.3	3.7	**54.3407**	0.756	1.4	ug/L	1410	Standard
	Sc	45	39518.2	3.2				ug/L	39543	Standard
	Ti	47	24690.4	0.9	**102.0722**	0.550	0.5	ug/L	39	Standard
	V	51	333814.0	1.2	**49.2501**	0.769	1.6	ug/L	1415	Standard
	Cr	52	345665.5	1.2	**50.0774**	0.577	1.2	ug/L	6946	Standard
	Cr	53	43291.8	1.9	**50.4406**	0.792	1.6	ug/L	328	Standard
	Mn	55	319563.2	0.6	**50.6110**	0.452	0.9	ug/L	823	Standard
	Co	59	337944.7	0.8	**49.4985**	0.535	1.1	ug/L	252	Standard
	Ni	60	91064.9	0.9	**49.6861**	0.418	0.8	ug/L	192	Standard
	Cu	65	92447.5	0.3	**50.0106**	0.369	0.7	ug/L	141	Standard
	Zn	66	53973.0	1.5	**49.8813**	0.957	1.9	ug/L	491	Standard
>	Ge	72	500217.2	0.4				ug/L	515087	Standard
	As	75	53742.2	1.5	**49.2630**	0.862	1.7	ug/L	20	Standard
	Se	82	5467.6	1.7	**48.7858**	1.017	2.1	ug/L	15	Standard
	Se-1	77	3788.8	1.0	**51.0702**	0.720	1.4	ug/L	48	Standard
>	Ga	71	51.7	5.6				mg/L	90	Standard
	Rb	85	820.0	10.6				ug/L	23	Standard
	Y	89	441206.7	0.2				ug/L	448510	Standard
>	Rh	103	33.3	8.7				ug/L	15	Standard
	Mo	98	251795.6	1.0	**94.4274**	1.474	1.6	ug/L	11	Standard
	Ag	107	307525.6	0.6	**46.4306**	0.349	0.8	ug/L	134	Standard
	Cd	111	100206.1	1.3	**47.5308**	1.177	2.5	mg/L	7	Standard
	Cd	114	244073.0	1.7	**47.6209**	0.821	1.7	ug/L	16	Standard
>	In	115	602825.2	1.4				ug/L	611117	Standard
	Sn	118	281047.8	1.1	**48.0844**	0.236	0.5	ug/L	518	Standard
	Sb	123	222239.4	0.2	**47.7427**	0.588	1.2	ug/L	127	Standard
	Ba	135	96470.5	0.6	**47.3499**	0.512	1.1	ug/L	45	Standard
	Ce	140	125.0	48.0				ug/L	298	Standard
>	Tb	159	921870.8	1.1				ug/L	977942	Standard
	Ho	165	11.7	65.5				ug/L	12	Standard
	Tl	203	392459.6	0.5	**49.4989**	0.479	1.0	ug/L	31	Standard
	Tl	205	352339.7	0.4	**51.1783**	0.770	1.5	ug/L	20	Standard
	Pb	206	250499.8	0.3	**49.1963**	0.428	0.9	ug/L	286	Standard
	Pb	207	226961.6	0.5	**49.4195**	0.337	0.7	ug/L	257	Standard
	Pb	208	926488.8	0.6	**52.1497**	0.850	1.6	ug/L	956	Standard
	U	238	343296.3	1.3	**50.9156**	0.444	0.9	ug/L	17	Standard
>	Bi	209	524204.3	1.1				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:59:47
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.4692**	18.124	173.1	mg/L	0	Standard
Mg	24	3737.1	0.3	**4.9245**	0.180	3.7	mg/L	32	Standard
K	39	1403.4	10.2	**4.6088**	0.536	11.6	mg/L	30	Standard
Ca	43	78.3	16.1	**6.5318**	1.619	24.8	mg/L	48	Standard
Fe	54	5604.3	1.6	**4.9425**	0.217	4.4	mg/L	159	Standard
Fe	57	1713.4	2.9	**5.1658**	0.187	3.6	mg/L	195	Standard
Sc-1	45	39518.2	3.2				mg/L	39543	Standard
Cl	35	33439.5	0.8				ug/L	27642	Standard
Kr	83	2.7	57.3				ug/L	2	Standard
Br	81	1523.4	5.0				ug/L	1647	Standard
P	31	18217.6	1.7				ug/L	15231	Standard
S	34	2641.9	8.4				ug/L	1845	Standard
Sr	88	48.3	11.9				ug/L	70	Standard
C	12	93.3	12.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	9.4	104.5				mg/L	12	Standard
Ho-1	165	11.7	65.5				mg/L	12	Standard
Er	166	13.3	114.6				mg/L	20	Standard
I	127	1510.1	17.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	103.188		
Al	27	108.681		
Sc	45			
Ti	47	102.072		
V	51	98.500		
Cr	52	100.155		
Cr	53			
Mn	55	101.222		
Co	59	98.997		
Ni	60	99.372		
Cu	65	100.021		
Zn	66	99.763		
Ge	72		97.113	
As	75	98.526		
Se	82	97.572		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:59:47
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	94.427		
Ag	107	92.861		
Cd	111	95.062		
Cd	114			
> In	115			98.643
Sn	118	96.169		
Sb	123	95.485		
Ba	135	94.700		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	98.998		
Tl	205			
Pb	206			
Pb	207			
Pb	208	104.299		
U	238	101.831		
> Bi	209			94.098
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:59:47
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 00:00:41
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70070.0	1.8				ug/L	81760	Standard
	Be	9	30.0	16.7	**0.0256**	0.006	25.2	ug/L	17	Standard
	Al	27	3812.2	30.7	**0.0144**	0.008	57.9	ug/L	1410	Standard
	Sc	45	40014.5	4.3				ug/L	39543	Standard
	Ti	47	26.7	12.1	**-0.0568**	0.013	23.6	ug/L	39	Standard
	V	51	1145.1	6.3	**-0.0403**	0.010	25.4	ug/L	1415	Standard
	Cr	52	6171.9	1.7	**-0.1619**	0.014	8.6	ug/L	6946	Standard
	Cr	53	411.7	3.9	**0.1320**	0.019	14.5	ug/L	328	Standard
	Mn	55	924.7	19.7	**-0.0044**	0.029	653.9	ug/L	823	Standard
	Co	59	261.3	12.2	**0.0062**	0.005	75.0	ug/L	252	Standard
	Ni	60	171.3	8.8	**-0.0131**	0.008	64.0	ug/L	192	Standard
	Cu	65	132.0	5.3	**-0.0207**	0.004	18.6	ug/L	141	Standard
	Zn	66	427.3	9.5	**-0.4339**	0.039	8.9	ug/L	491	Standard
>	**Ge**	72	510161.6	0.4				ug/L	515087	Standard
	As	75	-21.8	127.1	**0.0207**	0.025	120.4	ug/L	20	Standard
	Se	82	17.0	25.9	**0.0293**	0.038	129.6	ug/L	15	Standard
	Se-1	77	53.7	12.0	**0.0005**	0.088	17869.4	ug/L	48	Standard
>	**Ga**	71	35.0	28.6				mg/L	90	Standard
	Rb	85	33.3	60.6				ug/L	23	Standard
	Y	89	437929.1	1.0				ug/L	448510	Standard
>	**Rh**	103	6.7	43.3				ug/L	15	Standard
	Mo	98	269.2	32.9	**0.0838**	0.033	39.3	ug/L	11	Standard
	Ag	107	120.7	16.7	**0.0024**	0.003	124.0	ug/L	134	Standard
	Cd	111	21.8	34.3	**0.0077**	0.004	48.1	mg/L	7	Standard
	Cd	114	36.7	62.4	**0.0024**	0.005	190.6	ug/L	16	Standard
>	**In**	115	590666.4	0.9				ug/L	611117	Standard
	Sn	118	743.4	12.2	**0.0282**	0.017	60.4	ug/L	518	Standard
	Sb	123	532.5	22.4	**0.1040**	0.027	25.7	ug/L	127	Standard
	Ba	135	71.0	29.1	**0.0104**	0.010	96.8	ug/L	45	Standard
	Ce	140	91.7	65.8				ug/L	298	Standard
>	**Tb**	159	911246.7	0.8				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	81.7	41.2	**0.0055**	0.004	77.1	ug/L	31	Standard
	Tl	205	68.3	66.0	**0.0076**	0.007	85.7	ug/L	20	Standard
	Pb	206	297.7	10.9	**-0.0054**	0.007	123.4	ug/L	286	Standard
	Pb	207	283.7	16.0	**0.0035**	0.010	283.6	ug/L	257	Standard
	Pb	208	1155.4	11.3	**0.0043**	0.008	176.0	ug/L	956	Standard
	U	238	70.3	39.4	**0.0074**	0.004	55.5	ug/L	17	Standard
>	**Bi**	209	527005.8	1.2				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 00:02:58
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	25.0	20.0	**-0.0244**	0.005	21.9	mg/L	32	Standard
K	39	18.3	31.5	**-0.0799**	0.021	26.0	mg/L	30	Standard
Ca	43	16.7	17.3	**-2.8535**	0.318	11.1	mg/L	48	Standard
Fe	54	122.6	39.6	**-0.0261**	0.047	179.6	mg/L	159	Standard
Fe	57	230.0	17.0	**0.1451**	0.148	101.7	mg/L	195	Standard
> **Sc-1**	45	40014.5	4.3				mg/L	39543	Standard
Cl	35	33255.8	1.4				ug/L	27642	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	1523.4	4.7				ug/L	1647	Standard
P	31	17488.4	2.6				ug/L	15231	Standard
S	34	2358.5	3.1				ug/L	1845	Standard
Sr	88	70.0	25.8				ug/L	70	Standard
C	12	73.3	39.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.5	100.0				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	1368.4	18.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.044	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 00:02:58
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.654
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.601
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 00:02:58
Page 3

Approved: June 29, 2016

MassCal File Name

Mass Calibration File Name Default.tun
MassCal File Path C:\NexIONData\MassCal\Default.tun
Peak Search Window: 1.00

Sample Information

Sample Date/Time: Wednesday, June 29, 2016 09:50:03

Mass Calibration and Resolution

Analyte	E Mass	Meas Mass	Mass C DAC Val	Res DAC Value	Meas Peak W	Custom Res
Li	7.016	6.975	1323	2024	0.699	
Mg	23.985	23.975	4506	2020	0.705	
Co	58.933	58.975	11694	2022	0.697	
In	114.904	114.925	22865	2028	0.690	
U	238.050	238.025	47450	2043	0.688	

Relative Std. Dev.

Mass	Meas. Intens. RSD
5.525	9.254
5.575	1.459
5.625	2.417
5.675	3.442
5.725	2.231
5.775	2.296
5.825	3.784
5.875	3.197
5.925	2.400
5.975	1.087
6.025	1.670
6.075	3.521
6.125	2.277
6.175	2.559
6.225	20.588
6.275	74.689
6.325	108.327
6.375	83.853
6.425	69.722
6.475	32.808
6.525	10.642
6.575	8.985
6.625	1.429
6.675	0.941
6.725	0.857
6.775	3.333
6.825	0.822

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 1

Approved: July 04, 2016

6.875	1.277
6.925	1.899
6.975	1.561
7.025	1.195
7.075	1.059
7.125	1.545
7.175	1.743
7.225	3.248
7.275	8.142
7.325	57.701
7.375	91.287
7.425	51.602
7.475	55.902
7.525	100.000
7.575	69.869
7.625	81.441
7.675	81.441
7.725	70.711
7.775	55.902
7.825	59.266
7.875	49.793
7.925	104.583
7.975	98.543
8.025	94.786
8.075	223.607
8.125	81.441
8.175	84.265
8.225	61.237
8.275	81.312
8.325	81.441
8.375	91.287
8.425	82.402
8.475	22.822
22.525	149.071
22.575	52.705
22.625	46.771
22.675	48.627
22.725	39.087
22.775	18.109
22.825	22.048
22.875	36.422
22.925	32.723
22.975	82.685
23.025	39.239
23.075	25.529
23.125	43.064
23.175	21.317

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 2

Approved: July 04, 2016

23.225	22.232
23.275	15.408
23.325	42.127
23.375	37.172
23.425	24.296
23.475	14.668
23.525	3.807
23.575	1.593
23.625	2.258
23.675	1.180
23.725	1.542
23.775	1.943
23.825	1.185
23.875	1.595
23.925	1.324
23.975	1.055
24.025	1.809
24.075	1.103
24.125	1.249
24.175	0.652
24.225	1.203
24.275	3.614
24.325	39.316
24.375	52.355
24.425	20.738
24.475	5.001
24.525	5.067
24.575	0.930
24.625	1.241
24.675	0.972
24.725	1.552
24.775	0.581
24.825	0.736
24.875	1.684
24.925	0.952
24.975	1.395
25.025	0.762
25.075	1.498
25.125	1.418
25.175	1.145
25.225	6.288
25.275	44.605
25.325	50.990
25.375	73.823
25.425	57.334
25.475	7.010
57.525	8.193

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 3

Approved: July 04, 2016

57.575	3.609
57.625	2.367
57.675	3.747
57.725	3.869
57.775	3.226
57.825	1.406
57.875	2.220
57.925	2.049
57.975	2.397
58.025	3.045
58.075	0.897
58.125	1.397
58.175	1.608
58.225	1.481
58.275	5.244
58.325	6.996
58.375	14.816
58.425	39.994
58.475	18.390
58.525	7.862
58.575	1.876
58.625	1.684
58.675	0.757
58.725	2.431
58.775	2.139
58.825	2.293
58.875	1.987
58.925	2.134
58.975	1.760
59.025	2.893
59.075	1.855
59.125	4.126
59.175	3.545
59.225	2.980
59.275	6.918
59.325	57.198
59.375	79.057
59.425	39.123
59.475	32.983
59.525	7.337
59.575	10.704
59.625	4.888
59.675	5.230
59.725	3.575
59.775	2.279
59.825	3.495
59.875	0.859

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 4

Approved: July 04, 2016

59.925	2.164
59.975	4.187
60.025	3.749
60.075	1.767
60.125	4.373
60.175	1.581
60.225	1.647
60.275	29.768
60.325	42.377
60.375	71.261
60.425	71.261
60.475	69.722
113.525	10.800
113.575	6.260
113.625	1.672
113.675	1.646
113.725	1.420
113.775	2.281
113.825	1.446
113.875	0.820
113.925	1.607
113.975	1.407
114.025	1.046
114.075	2.403
114.125	2.090
114.175	3.331
114.225	5.316
114.275	12.354
114.325	37.872
114.375	26.082
114.425	15.800
114.475	3.831
114.525	4.062
114.575	1.808
114.625	2.980
114.675	3.272
114.725	2.895
114.775	1.904
114.825	1.941
114.875	1.809
114.925	0.906
114.975	0.414
115.025	2.617
115.075	0.808
115.125	1.618
115.175	2.428
115.225	3.339

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 5

Approved: July 04, 2016

115.275	5.459
115.325	14.203
115.375	69.869
115.425	66.254
115.475	26.966
115.525	18.235
115.575	7.589
115.625	3.788
115.675	2.661
115.725	3.661
115.775	2.867
115.825	3.952
115.875	6.794
115.925	4.558
115.975	5.421
116.025	1.319
116.075	4.928
116.125	7.259
116.175	4.689
116.225	9.289
116.275	22.491
116.325	26.307
116.375	26.146
116.425	67.711
116.475	40.825
236.525	
236.575	13.258
236.625	39.219
236.675	66.333
236.725	37.815
236.775	20.039
236.825	39.510
236.875	22.934
236.925	27.407
236.975	8.537
237.025	28.853
237.075	45.928
237.125	41.937
237.175	36.377
237.225	26.058
237.275	27.285
237.325	43.196
237.375	45.018
237.425	23.111
237.475	20.908
237.525	26.413
237.575	10.671

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 6

Approved: July 04, 2016

237.625	6.558
237.675	4.584
237.725	2.707
237.775	1.019
237.825	0.867
237.875	0.837
237.925	1.315
237.975	0.556
238.025	1.099
238.075	0.473
238.125	0.980
238.175	0.943
238.225	1.564
238.275	1.714
238.325	1.785
238.375	2.057
238.425	2.571
238.475	6.163
238.525	10.798
238.575	23.407
238.625	35.493
238.675	33.171
238.725	15.721
238.775	9.421
238.825	23.516
238.875	27.721
238.925	14.417
238.975	30.010
239.025	24.993
239.075	29.933
239.125	17.678
239.175	38.433
239.225	46.981
239.275	15.215
239.325	48.601
239.375	50.659
239.425	23.570
239.475	38.955

Approved: July 04, 2016

SmartTune Wizard - Summary

```
Optimization Summary

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Start Time: 6/29/2016 10:01:57 AM
End Time: 6/29/2016 10:04:08 AM

Daily Performance Check - [Passed]  Optimum value(s): N/A
        Obtained Intensity (Be 9.0122): 4651.78
        Obtained Intensity (Mg 23.985): 162439.88
        Obtained Intensity (In 114.904): 66202.13
        Obtained Intensity (U 238.05): 140216.56
        Obtained Intensity (Bkgd 220): 0.33
        Obtained Formula (CeO 155.9 / Ce 139.905): 0.018 (=3640.50 / 205153.73)
        Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.004 (=909.89 / 205153.73)
```


Approved: July 04, 2016

SmartTune Wizard - Details

```
Optimization Details

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Optimization Status

Start Time: 6/29/2016 10:01:57 AM

Daily Performance Check
        Optimization Settings:
                Method: C:\NexIONData\Method\ESI Daily Performance.mth.
                Intensity Criterion: Be 9.0122 > 2000
                Intensity Criterion: Mg 23.985 > 15000
                Intensity Criterion: In 114.904 > 40000
                Intensity Criterion: U 238.05 > 30000
                Intensity Criterion: Bkgd 220 <= 5
                Formula Criterion: CeO 155.9 / Ce 139.905 <= 0.025
                Formula Criterion: Ce++ 69.9527 / Ce 139.905 <= 0.03

        Optimization Results:
        Initial Try
                Obtained Intensity (Be 9.0122): 4651.78
                Obtained Intensity (Mg 23.985): 162439.88
                Obtained Intensity (In 114.904): 66202.13
                Obtained Intensity (U 238.05): 140216.56
                Obtained Intensity (Bkgd 220): 0.33
                Obtained Formula (CeO 155.9 / Ce 139.905): 0.018 (=3640.50 / 205153.73)
                Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.004 (=909.89 / 205153.73)
[Passed]  Optimum value(s): N/A

End Time: 6/29/2016 10:04:08 AM
```


Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Blank
Sample Date/Time: Wednesday, June 29, 2016 11:22:34
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	74698.8	3.1				ug/L		Standard
	Be	9	8.3	124.9				ug/L		Standard
	Al	27	1493.4	9.1				ug/L		Standard
	Sc	45	61029.4	2.9				ug/L		Standard
	Ti	47	35.0	12.5				ug/L		Standard
	V	51	1434.8	4.3				ug/L		Standard
	Cr	52	7634.6	2.5				ug/L		Standard
	Cr	53	363.3	0.8				ug/L		Standard
	Mn	55	747.0	1.6				ug/L		Standard
	Co	59	233.7	7.1				ug/L		Standard
	Ni	60	214.7	3.5				ug/L		Standard
	Cu	65	104.0	12.0				ug/L		Standard
	Zn	66	363.0	3.8				ug/L		Standard
>	Ge	72	495142.3	1.3				ug/L		Standard
	As	75	-49.8	7.8				ug/L		Standard
	Se	82	13.7	23.9				ug/L		Standard
	Se-1	77	42.3	9.8				ug/L		Standard
>	Ga	71	41.7	25.0				mg/L		Standard
	Rb	85	16.7	91.7				ug/L		Standard
	Y	89	417098.4	0.9				ug/L		Standard
>	Rh	103	13.3	86.6				ug/L		Standard
	Mo	98	10.7	10.9				ug/L		Standard
	Ag	107	68.0	6.4				ug/L		Standard
	Cd	111	2.3	24.9				mg/L		Standard
	Cd	114	11.7	106.5				ug/L		Standard
>	In	115	587441.1	0.5				ug/L		Standard
	Sn	118	496.7	9.7				ug/L		Standard
	Sb	123	77.7	34.7				ug/L		Standard
	Ba	135	24.3	19.4				ug/L		Standard
	Ce	140	25.0	20.0				ug/L		Standard
>	Tb	159	939701.0	1.8				ug/L		Standard
	Ho	165	10.0	50.0				ug/L		Standard
	Tl	203	4.0	25.0				ug/L		Standard
	Tl	205	1.7	173.2				ug/L		Standard
	Pb	206	239.7	3.6				ug/L		Standard
	Pb	207	197.7	3.9				ug/L		Standard
	Pb	208	883.3	3.3				ug/L		Standard
	U	238	3.3	17.3				ug/L		Standard
>	Bi	209	523984.5	1.3				ug/L		Standard

Sample ID: Blank
Report Date/Time: Wednesday, June 29, 2016 11:24:51
Page 1

Approved: July 04, 2016

	Na	23	0.0		mg/L	Standard
	Mg	24	36.7	55.1	mg/L	Standard
	K	39	38.3	52.7	mg/L	Standard
	Ca	43	10.0	50.0	mg/L	Standard
	Fe	54	179.1	14.9	mg/L	Standard
	Fe	57	335.0	10.3	mg/L	Standard
>	**Sc-1**	45	61029.4	2.9	mg/L	Standard
	Cl	35	20194.8	0.8	ug/L	Standard
	Kr	83	4.3	70.5	ug/L	Standard
	Br	81	1546.7	2.4	ug/L	Standard
	P	31	15286.0	3.7	ug/L	Standard
	S	34	1571.7	2.7	ug/L	Standard
	Sr	88	73.3	3.9	ug/L	Standard
	C	12	63.3	24.1	mg/L	Standard
	N	14	0.0		mg/L	Standard
	Hg	202	3.3	173.2	mg/L	Standard
	Dy	164	6.0	187.1	mg/L	Standard
	Ho-1	165	10.0	50.0	mg/L	Standard
	Er	166	13.3	86.6	mg/L	Standard
	I	127	1391.7	17.5	mg/L	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Blank
Report Date/Time: Wednesday, June 29, 2016 11:24:51
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Blank
Report Date/Time: Wednesday, June 29, 2016 11:24:51
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 1
Sample Date/Time: Wednesday, June 29, 2016 11:25:45
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75658.5	2.0				ug/L	74699	Standard
	Be	9	1.7	173.2				ug/L	8	Standard
	Al	27	1555.1	3.1				ug/L	1493	Standard
	Sc	45	60145.9	2.9				ug/L	61029	Standard
	Ti	47	32.7	4.7				ug/L	35	Standard
	V	51	1434.9	4.0				ug/L	1435	Standard
	Cr	52	7544.5	2.5				ug/L	7635	Standard
	Cr	53	365.0	18.0				ug/L	363	Standard
	Mn	55	783.7	0.8				ug/L	747	Standard
	Co	59	239.0	2.9				ug/L	234	Standard
	Ni	60	220.7	1.8				ug/L	215	Standard
	Cu	65	103.0	8.3				ug/L	104	Standard
	Zn	66	365.0	4.1				ug/L	363	Standard
>	**Ge**	72	495351.7	2.1				ug/L	495142	Standard
	As	75	-40.0	8.7				ug/L	-50	Standard
	Se	82	19.1	38.4				ug/L	14	Standard
	Se-1	77	47.3	15.6				ug/L	42	Standard
>	**Ga**	71	53.3	47.2				mg/L	42	Standard
	Rb	85	15.0	88.2				ug/L	17	Standard
	Y	89	427717.4	1.6				ug/L	417098	Standard
>	**Rh**	103	8.3	34.6				ug/L	13	Standard
	Mo	98	11.1	31.8				ug/L	11	Standard
	Ag	107	86.3	19.8				ug/L	68	Standard
	Cd	111	1.3	43.8				mg/L	2	Standard
	Cd	114	3.6	152.7				ug/L	12	Standard
>	**In**	115	599155.0	1.0				ug/L	587441	Standard
	Sn	118	486.7	3.3				ug/L	497	Standard
	Sb	123	28.2	25.8				ug/L	78	Standard
	Ba	135	25.3	13.9				ug/L	24	Standard
	Ce	140	31.7	59.8				ug/L	25	Standard
>	**Tb**	159	985563.2	0.5				ug/L	939701	Standard
	Ho	165	6.7	86.6				ug/L	10	Standard
	Tl	203	1.7	91.7				ug/L	4	Standard
	Tl	205	0.0					ug/L	2	Standard
	Pb	206	249.7	4.0				ug/L	240	Standard
	Pb	207	216.0	2.0				ug/L	198	Standard
	Pb	208	873.0	4.1				ug/L	883	Standard
	U	238	3.3	34.6				ug/L	3	Standard
>	**Bi**	209	535312.9	0.6				ug/L	523984	Standard

Sample ID: Standard 1
Report Date/Time: Wednesday, June 29, 2016 11:28:02
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	mg/L	0	Standard
Mg	24	31.7	18.2	mg/L	37	Standard
K	39	26.7	28.6	mg/L	38	Standard
Ca	43	35.0	24.7	mg/L	10	Standard
Fe	54	130.8	15.0	mg/L	179	Standard
Fe	57	305.0	2.8	mg/L	335	Standard
Sc-1	45	60145.9	2.9	mg/L	61029	Standard
Cl	35	20341.6	1.8	ug/L	20195	Standard
Kr	83	3.0	33.3	ug/L	4	Standard
Br	81	1546.7	10.4	ug/L	1547	Standard
P	31	15145.9	1.8	ug/L	15286	Standard
S	34	1610.1	6.1	ug/L	1572	Standard
Sr	88	86.7	14.5	ug/L	73	Standard
C	12	60.0	28.9	mg/L	63	Standard
N	14	0.0		mg/L	0	Standard
Hg	202	0.0		mg/L	3	Standard
Dy	164	5.7	108.5	mg/L	6	Standard
Ho-1	165	6.7	86.6	mg/L	10	Standard
Er	166	20.0	50.0	mg/L	13	Standard
I	127	1368.4	7.8	mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72			
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: Standard 1
Report Date/Time: Wednesday, June 29, 2016 11:28:02
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 1
Report Date/Time: Wednesday, June 29, 2016 11:28:02
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 2
Sample Date/Time: Wednesday, June 29, 2016 11:28:57
Number of Replicates: 3
Autosampler Position: 2
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73081.0	3.5				ug/L	74699	Standard
	Be	9	36.7	56.8				ug/L	8	Standard
	Al	27	9102.7	2.7				ug/L	1493	Standard
	Sc	45	58226.8	3.4				ug/L	61029	Standard
	Ti	47	79.0	2.2				ug/L	35	Standard
	V	51	1848.7	2.3				ug/L	1435	Standard
	Cr	52	7933.1	1.4				ug/L	7635	Standard
	Cr	53	350.0	15.1				ug/L	363	Standard
	Mn	55	1375.4	4.6				ug/L	747	Standard
	Co	59	574.7	5.8				ug/L	234	Standard
	Ni	60	324.3	7.6				ug/L	215	Standard
	Cu	65	285.0	4.9				ug/L	104	Standard
	Zn	66	549.3	0.4				ug/L	363	Standard
>	**Ge**	72	493663.5	1.5				ug/L	495142	Standard
	As	75	-10.4	282.1				ug/L	-50	Standard
	Se	82	20.6	18.5				ug/L	14	Standard
	Se-1	77	49.0	19.5				ug/L	42	Standard
>	**Ga**	71	58.3	27.6				mg/L	42	Standard
	Rb	85	25.0	60.0				ug/L	17	Standard
	Y	89	437555.0	0.8				ug/L	417098	Standard
>	**Rh**	103	25.0	34.6				ug/L	13	Standard
	Mo	98	283.4	4.8				ug/L	11	Standard
	Ag	107	412.3	3.2				ug/L	68	Standard
	Cd	111	121.8	9.1				mg/L	2	Standard
	Cd	114	268.9	5.0				ug/L	12	Standard
>	**In**	115	590951.3	0.9				ug/L	587441	Standard
	Sn	118	850.0	6.4				ug/L	497	Standard
	Sb	123	254.5	2.7				ug/L	78	Standard
	Ba	135	136.0	15.7				ug/L	24	Standard
	Ce	140	18.3	56.8				ug/L	25	Standard
>	**Tb**	159	941723.8	1.9				ug/L	939701	Standard
	Ho	165	3.3	86.6				ug/L	10	Standard
	Tl	203	417.3	5.5				ug/L	4	Standard
	Tl	205	348.3	7.4				ug/L	2	Standard
	Pb	206	633.0	0.8				ug/L	240	Standard
	Pb	207	557.0	2.2				ug/L	198	Standard
	Pb	208	2275.4	1.7				ug/L	883	Standard
	U	238	366.3	0.7				ug/L	3	Standard
>	**Bi**	209	512414.5	1.5				ug/L	523984	Standard

Sample ID: Standard 2
Report Date/Time: Wednesday, June 29, 2016 11:31:14
Page 1

Approved: July 04, 2016

	Na	23	0.0		mg/L	0	Standard
	Mg	24	51.7	24.4	mg/L	37	Standard
	K	39	51.7	24.4	mg/L	38	Standard
	Ca	43	38.3	19.9	mg/L	10	Standard
	Fe	54	160.9	17.5	mg/L	179	Standard
	Fe	57	303.3	9.1	mg/L	335	Standard
>	**Sc-1**	45	58226.8	3.4	mg/L	61029	Standard
	Cl	35	20887.7	1.1	ug/L	20195	Standard
	Kr	83	2.3	107.9	ug/L	4	Standard
	Br	81	1533.4	9.8	ug/L	1547	Standard
	P	31	14952.4	2.3	ug/L	15286	Standard
	S	34	1701.8	4.5	ug/L	1572	Standard
	Sr	88	66.7	31.2	ug/L	73	Standard
	C	12	46.7	65.5	mg/L	63	Standard
	N	14	0.0		mg/L	0	Standard
	Hg	202	0.0		mg/L	3	Standard
	Dy	164	12.7	89.9	mg/L	6	Standard
	Ho-1	165	3.3	86.6	mg/L	10	Standard
	Er	166	13.3	43.3	mg/L	13	Standard
	I	127	1343.4	21.4	mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Standard 2
Report Date/Time: Wednesday, June 29, 2016 11:31:14
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 2
Report Date/Time: Wednesday, June 29, 2016 11:31:14
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 3
Sample Date/Time: Wednesday, June 29, 2016 11:32:08
Number of Replicates: 3
Autosampler Position: 3
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70836.9	0.7				ug/L	74699	Standard
	Be	9	41495.2	4.9	**50.0000**	2.291	4.6	ug/L	8	Standard
	Al	27	6913642.3	1.4	**50.0000**	0.333	0.7	ug/L	1493	Standard
	Sc	45	58109.6	1.6				ug/L	61029	Standard
	Ti	47	38024.7	0.9	**100.0000**	1.896	1.9	ug/L	35	Standard
	V	51	385028.2	1.3	**50.0000**	1.023	2.0	ug/L	1435	Standard
	Cr	52	378864.5	0.8	**50.0000**	0.813	1.6	ug/L	7635	Standard
	Cr	53	46780.4	0.7	**50.0000**	0.853	1.7	ug/L	363	Standard
	Mn	55	354877.0	1.3	**50.0000**	1.137	2.3	ug/L	747	Standard
	Co	59	336932.8	1.2	**50.0000**	1.027	2.1	ug/L	234	Standard
	Ni	60	94633.6	0.8	**50.0000**	0.638	1.3	ug/L	215	Standard
	Cu	65	94252.3	1.0	**50.0000**	0.913	1.8	ug/L	104	Standard
	Zn	66	55983.3	1.5	**50.0000**	1.110	2.2	ug/L	363	Standard
>	**Ge**	72	497274.4	1.0				ug/L	495142	Standard
	As	75	56952.3	0.9	**50.0000**	0.907	1.8	ug/L	-50	Standard
	Se	82	5614.6	2.5	**50.0000**	1.630	3.3	ug/L	14	Standard
	Se-1	77	3760.8	2.5	**50.0000**	1.757	3.5	ug/L	42	Standard
>	**Ga**	71	60.0	22.0				mg/L	42	Standard
	Rb	85	735.0	7.4				ug/L	17	Standard
	Y	89	413691.0	2.8				ug/L	417098	Standard
>	**Rh**	103	28.3	27.0				ug/L	13	Standard
	Mo	98	261684.0	1.3	**100.0000**	1.411	1.4	ug/L	11	Standard
	Ag	107	317580.7	1.1	**50.0000**	0.030	0.1	ug/L	68	Standard
	Cd	111	102024.7	1.3	**50.0000**	0.465	0.9	mg/L	2	Standard
	Cd	114	251691.0	1.5	**50.0000**	1.089	2.2	ug/L	12	Standard
>	**In**	115	567339.3	1.1				ug/L	587441	Standard
	Sn	118	284919.2	1.1	**50.0000**	0.694	1.4	ug/L	497	Standard
	Sb	123	226852.6	0.2	**50.0000**	0.610	1.2	ug/L	78	Standard
	Ba	135	99575.8	0.1	**50.0000**	0.595	1.2	ug/L	24	Standard
	Ce	140	83.3	34.6				ug/L	25	Standard
>	**Tb**	159	932165.5	0.9				ug/L	939701	Standard
	Ho	165	13.3	21.7				ug/L	10	Standard
	Tl	203	405993.6	0.4	**50.0000**	0.749	1.5	ug/L	4	Standard
	Tl	205	349857.0	1.5	**50.0000**	0.282	0.6	ug/L	2	Standard
	Pb	206	265201.9	0.6	**50.0000**	0.817	1.6	ug/L	240	Standard
	Pb	207	235754.7	0.6	**50.0000**	0.829	1.7	ug/L	198	Standard
	Pb	208	926203.5	0.6	**50.0000**	0.626	1.3	ug/L	883	Standard
	U	238	337413.2	1.0	**50.0000**	0.543	1.1	ug/L	3	Standard
>	**Bi**	209	511781.8	1.3				ug/L	523984	Standard

Sample ID: Standard 3
Report Date/Time: Wednesday, June 29, 2016 11:34:25
Page 1

Approved: July 04, 2016

Na	23	0.0					mg/L	0	Standard
Mg	24	3782.1	6.2	**5.0000**	0.388	7.8	mg/L	37	Standard
K	39	2271.8	5.5	**5.0000**	0.254	5.1	mg/L	38	Standard
Ca	43	120.0	23.2	**5.0000**	1.671	33.4	mg/L	10	Standard
Fe	54	5877.0	6.9	**5.0000**	0.418	8.4	mg/L	179	Standard
Fe	57	1703.4	4.4	**5.0000**	0.228	4.6	mg/L	335	Standard
> **Sc-1**	45	58109.6	1.6				mg/L	61029	Standard
Cl	35	22061.4	1.9				ug/L	20195	Standard
Kr	83	2.7	94.4				ug/L	4	Standard
Br	81	1433.4	2.5				ug/L	1547	Standard
P	31	17349.9	2.8				ug/L	15286	Standard
S	34	2026.8	6.8				ug/L	1572	Standard
Sr	88	76.7	26.4				ug/L	73	Standard
C	12	43.3	48.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	33.0	63.6				mg/L	6	Standard
Ho-1	165	13.3	21.7				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1500.1	14.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72			
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: Standard 3
Report Date/Time: Wednesday, June 29, 2016 11:34:25
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 4
Sample Date/Time: Wednesday, June 29, 2016 11:35:20
Number of Replicates: 3
Autosampler Position: 4
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	70562.3	0.6				ug/L	74699	Standard
	Be	9	83650.3	2.2	**100.5872**	1.689	1.7	ug/L	8	Standard
	Al	27	13996690.0	1.6	**100.8115**	1.548	1.5	ug/L	1493	Standard
	Sc	45	58977.9	1.4				ug/L	61029	Standard
	Ti	47	75710.7	1.4	**200.4046**	3.310	1.7	ug/L	35	Standard
	V	51	745374.0	1.7	**98.8583**	0.465	0.5	ug/L	1435	Standard
	Cr	52	747654.3	1.6	**100.2327**	0.593	0.6	ug/L	7635	Standard
	Cr	53	92243.3	0.8	**99.8509**	0.489	0.5	ug/L	363	Standard
	Mn	55	686210.3	1.7	**98.7774**	1.044	1.1	ug/L	747	Standard
	Co	59	647972.6	0.9	**98.4550**	0.479	0.5	ug/L	234	Standard
	Ni	60	186003.6	1.6	**99.5867**	0.367	0.4	ug/L	215	Standard
	Cu	65	184421.8	1.6	**99.3502**	0.739	0.7	ug/L	104	Standard
	Zn	66	108058.7	1.7	**98.8348**	0.723	0.7	ug/L	363	Standard
>	Ge	72	493312.7	1.3				ug/L	495142	Standard
	As	75	109996.7	2.4	**98.6160**	1.152	1.2	ug/L	-50	Standard
	Se	82	10772.1	0.2	**98.3851**	1.295	1.3	ug/L	14	Standard
	Se-1	77	7226.7	2.3	**98.6814**	1.085	1.1	ug/L	42	Standard
>	Ga	71	88.3	19.9				mg/L	42	Standard
	Rb	85	1528.4	8.2				ug/L	17	Standard
	Y	89	411133.1	2.0				ug/L	417098	Standard
>	Rh	103	50.0	20.0				ug/L	13	Standard
	Mo	98	506691.6	2.0	**196.6866**	6.239	3.2	ug/L	11	Standard
	Ag	107	624459.9	0.9	**99.1146**	1.822	1.8	ug/L	68	Standard
	Cd	111	200021.3	0.6	**98.9639**	1.657	1.7	mg/L	2	Standard
	Cd	114	487602.9	1.3	**98.3617**	2.136	2.2	ug/L	12	Standard
>	In	115	567890.6	1.1				ug/L	587441	Standard
	Sn	118	560558.1	0.7	**99.1782**	1.388	1.4	ug/L	497	Standard
	Sb	123	448476.6	1.3	**99.3779**	2.340	2.4	ug/L	78	Standard
	Ba	135	196432.8	0.6	**99.2744**	1.775	1.8	ug/L	24	Standard
	Ce	140	161.7	46.3				ug/L	25	Standard
>	Tb	159	938936.1	1.2				ug/L	939701	Standard
	Ho	165	10.0					ug/L	10	Standard
	Tl	203	803908.2	0.6	**99.1382**	1.002	1.0	ug/L	4	Standard
	Tl	205	702848.2	0.9	**99.8673**	1.348	1.3	ug/L	2	Standard
	Pb	206	517801.9	1.0	**98.4705**	1.644	1.7	ug/L	240	Standard
	Pb	207	459764.9	1.8	**98.4124**	2.372	2.4	ug/L	198	Standard
	Pb	208	1827348.6	0.4	**98.9940**	0.966	1.0	ug/L	883	Standard
	U	238	674609.9	0.9	**99.6270**	1.321	1.3	ug/L	3	Standard
>	Bi	209	515462.4	0.6				ug/L	523984	Standard

Sample ID: Standard 4
Report Date/Time: Wednesday, June 29, 2016 11:37:37
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.9850**	34.606	173.2	mg/L	0	Standard
Mg	24	7435.2	2.2	**9.8676**	0.257	2.6	mg/L	37	Standard
K	39	4282.3	4.3	**9.6823**	0.481	5.0	mg/L	38	Standard
Ca	43	178.3	18.2	**9.1297**	2.109	23.1	mg/L	10	Standard
Fe	54	11328.6	1.8	**9.8033**	0.307	3.1	mg/L	179	Standard
Fe	57	3145.3	5.1	**9.9880**	0.532	5.3	mg/L	335	Standard
Sc-1	45	58977.9	1.4				mg/L	61029	Standard
Cl	35	21244.2	2.5				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	1583.4	3.9				ug/L	1547	Standard
P	31	17550.2	2.6				ug/L	15286	Standard
S	34	2090.1	3.5				ug/L	1572	Standard
Sr	88	83.3	42.6				ug/L	73	Standard
C	12	56.7	62.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	22.9	66.8				mg/L	6	Standard
Ho-1	165	10.0					mg/L	10	Standard
Er	166	10.0					mg/L	13	Standard
I	127	1638.4	19.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72			
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: Standard 4
Report Date/Time: Wednesday, June 29, 2016 11:37:37
Page 2

Approved: July 04, 2016

	Rb	85
	Y	89
>	Rh	103
	Mo	98
	Ag	107
	Cd	111
	Cd	114
>	In	115
	Sn	118
	Sb	123
	Ba	135
	Ce	140
>	Tb	159
	Ho	165
	Tl	203
	Tl	205
	Pb	206
	Pb	207
	Pb	208
	U	238
>	Bi	209
	Na	23
	Mg	24
	K	39
	Ca	43
	Fe	54
	Fe	57
>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Corr. Coef.	Na	23	Correlation coefficient < 0.998
Corr. Coef.	Ca	43	Correlation coefficient < 0.998

Sample ID: Standard 4
Report Date/Time: Wednesday, June 29, 2016 11:37:37
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 1
Sample Date/Time: Wednesday, June 29, 2016 11:38:33
Number of Replicates: 3
Autosampler Position: 201
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	69383.6	2.2				ug/L	74699	Standard
	Be	9	41665.5	1.0	**50.9761**	1.262	2.5	ug/L	8	Standard
	Al	27	7007266.1	0.7	**51.3351**	1.064	2.1	ug/L	1493	Standard
	Sc	45	57040.5	0.5				ug/L	61029	Standard
	Ti	47	39782.2	0.5	**104.1738**	1.427	1.4	ug/L	35	Standard
	V	51	390724.1	1.4	**51.2089**	1.334	2.6	ug/L	1435	Standard
	Cr	52	380689.4	0.9	**50.0149**	1.033	2.1	ug/L	7635	Standard
	Cr	53	46596.6	2.0	**49.7615**	1.193	2.4	ug/L	363	Standard
	Mn	55	364452.5	1.9	**51.8606**	1.232	2.4	ug/L	747	Standard
	Co	59	335357.4	2.0	**50.4145**	0.943	1.9	ug/L	234	Standard
	Ni	60	95528.2	0.6	**50.5685**	0.768	1.5	ug/L	215	Standard
	Cu	65	95355.8	0.8	**50.7995**	0.891	1.8	ug/L	104	Standard
	Zn	66	57795.4	0.9	**52.1095**	0.303	0.6	ug/L	363	Standard
>	Ge	72	498415.5	1.2				ug/L	495142	Standard
	As	75	57290.2	0.4	**50.8755**	0.674	1.3	ug/L	-50	Standard
	Se	82	5486.5	1.1	**49.5278**	0.838	1.7	ug/L	14	Standard
	Se-1	77	3797.8	2.0	**51.0456**	1.569	3.1	ug/L	42	Standard
>	Ga	71	88.3	43.2				mg/L	42	Standard
	Rb	85	576.7	2.8				ug/L	17	Standard
	Y	89	413113.3	1.2				ug/L	417098	Standard
>	Rh	103	26.7	92.5				ug/L	13	Standard
	Mo	98	275236.9	1.8	**107.0386**	1.474	1.4	ug/L	11	Standard
	Ag	107	320241.4	0.9	**50.9276**	0.747	1.5	ug/L	68	Standard
	Cd	111	101502.5	0.4	**50.3185**	0.332	0.7	mg/L	2	Standard
	Cd	114	248344.9	2.5	**50.2016**	1.511	3.0	ug/L	12	Standard
>	In	115	566676.1	0.5				ug/L	587441	Standard
	Sn	118	310270.4	1.9	**54.9671**	1.097	2.0	ug/L	497	Standard
	Sb	123	224427.6	0.4	**49.8273**	0.118	0.2	ug/L	78	Standard
	Ba	135	98607.5	0.6	**49.9289**	0.574	1.1	ug/L	24	Standard
	Ce	140	138.3	11.0				ug/L	25	Standard
>	Tb	159	931841.4	1.3				ug/L	939701	Standard
	Ho	165	10.0	86.6				ug/L	10	Standard
	Tl	203	410286.1	0.9	**50.6578**	0.710	1.4	ug/L	4	Standard
	Tl	205	350970.1	0.3	**49.9294**	0.473	0.9	ug/L	2	Standard
	Pb	206	272426.4	0.6	**51.8371**	0.812	1.6	ug/L	240	Standard
	Pb	207	234103.7	0.5	**50.1331**	0.525	1.0	ug/L	198	Standard
	Pb	208	940824.8	0.3	**50.9948**	0.649	1.3	ug/L	883	Standard
	U	238	333596.8	0.3	**49.3228**	0.442	0.9	ug/L	3	Standard
>	Bi	209	514849.0	1.0				ug/L	523984	Standard

Sample ID: QC Std 1
Report Date/Time: Wednesday, June 29, 2016 11:40:50
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	3822.1	2.3	**5.2142**	0.148	2.8	mg/L	37	Standard
K	39	2126.8	6.1	**4.9163**	0.328	6.7	mg/L	38	Standard
Ca	43	118.3	24.4	**5.4681**	1.919	35.1	mg/L	10	Standard
Fe	54	5948.1	0.7	**5.2575**	0.061	1.2	mg/L	179	Standard
Fe	57	1791.8	1.9	**5.4420**	0.142	2.6	mg/L	335	Standard
> **Sc-1**	45	57040.5	0.5				mg/L	61029	Standard
Cl	35	20895.1	0.7				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	1520.1	6.3				ug/L	1547	Standard
P	31	17665.3	1.6				ug/L	15286	Standard
S	34	2033.5	1.7				ug/L	1572	Standard
Sr	88	80.0	10.8				ug/L	73	Standard
C	12	70.0	24.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	23.0	24.5				mg/L	6	Standard
Ho-1	165	10.0	86.6				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1353.4	28.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	101.952		
Al	27	102.670		
Sc	45			
Ti	47	104.174		
V	51	102.418		
Cr	52	100.030		
Cr	53			
Mn	55	103.721		
Co	59	100.829		
Ni	60	101.137		
Cu	65	101.599		
Zn	66	104.219		
> Ge	72		100.661	
As	75	101.751		
Se	82	99.056		
Se-1	77			
> Ga	71			

Sample ID: QC Std 1
Report Date/Time: Wednesday, June 29, 2016 11:40:50
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	107.039		
Ag	107	101.855		
Cd	111	100.637		
Cd	114			
> In	115		96.465	
Sn	118	109.934		
Sb	123	99.655		
Ba	135	99.858		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.316		
Tl	205			
Pb	206	103.674		
Pb	207	100.266		
Pb	208	101.990		
U	238	98.646		
> Bi	209		98.257	
Na	23	0.100		
Mg	24	104.283		
K	39	98.326		
Ca	43	109.361		
Fe	54	105.151		
Fe	57	108.840		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 1	Na	23	
QC Std 1	Ca	43	

Sample ID: QC Std 1
Report Date/Time: Wednesday, June 29, 2016 11:40:50
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 2
Sample Date/Time: Wednesday, June 29, 2016 11:41:46
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70378.1	2.5				ug/L	74699	Standard
	Be	9	20.0	25.0	**0.0316**	0.005	17.2	ug/L	8	Standard
	Al	27	2221.8	11.1	**0.0027**	0.002	62.7	ug/L	1493	Standard
	Sc	45	58138.1	3.2				ug/L	61029	Standard
	Ti	47	47.0	20.3	**0.0161**	0.026	158.4	ug/L	35	Standard
	V	51	1216.9	5.3	**-0.0330**	0.008	24.4	ug/L	1435	Standard
	Cr	52	6478.4	1.5	**-0.1416**	0.009	6.0	ug/L	7635	Standard
	Cr	53	358.3	10.7	**0.0617**	0.046	74.0	ug/L	363	Standard
	Mn	55	785.0	2.1	**-0.0343**	0.003	7.5	ug/L	747	Standard
	Co	59	292.7	8.2	**0.0075**	0.004	50.2	ug/L	234	Standard
	Ni	60	209.7	4.3	**-0.0109**	0.006	51.3	ug/L	215	Standard
	Cu	65	121.0	5.8	**-0.0381**	0.003	8.4	ug/L	104	Standard
	Zn	66	398.3	6.4	**-0.0865**	0.023	26.3	ug/L	363	Standard
>	**Ge**	72	490643.1	1.0				ug/L	495142	Standard
	As	75	-2.3	851.2	**0.0577**	0.017	30.2	ug/L	-50	Standard
	Se	82	15.9	39.8	**0.0087**	0.059	680.3	ug/L	14	Standard
	Se-1	77	51.3	11.1	**0.0869**	0.077	88.1	ug/L	42	Standard
>	**Ga**	71	46.7	16.4				mg/L	42	Standard
	Rb	85	33.3	45.8				ug/L	17	Standard
	Y	89	424498.7	1.6				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	405.0	23.6	**0.1487**	0.033	22.3	ug/L	11	Standard
	Ag	107	128.3	5.2	**0.0071**	0.001	8.6	ug/L	68	Standard
	Cd	111	15.0	38.7	**-0.0006**	0.003	474.1	mg/L	2	Standard
	Cd	114	47.2	8.1	**0.0073**	0.001	11.6	ug/L	12	Standard
>	**In**	115	576512.9	2.1				ug/L	587441	Standard
	Sn	118	851.7	19.9	**0.0543**	0.032	58.8	ug/L	497	Standard
	Sb	123	1550.0	28.1	**0.3352**	0.100	29.7	ug/L	78	Standard
	Ba	135	36.7	23.2	**0.0022**	0.004	189.8	ug/L	24	Standard
	Ce	140	21.7	93.3				ug/L	25	Standard
>	**Tb**	159	924023.3	3.2				ug/L	939701	Standard
	Ho	165	10.0	100.0				ug/L	10	Standard
	Tl	203	60.7	18.1	**0.0057**	0.001	22.9	ug/L	4	Standard
	Tl	205	61.7	61.4	**0.0089**	0.005	59.3	ug/L	2	Standard
	Pb	206	271.3	7.0	**-0.0198**	0.004	19.8	ug/L	240	Standard
	Pb	207	237.3	7.1	**-0.0194**	0.004	21.0	ug/L	198	Standard
	Pb	208	1028.3	6.0	**-0.0187**	0.003	15.4	ug/L	883	Standard
	U	238	47.7	28.5	**0.0026**	0.002	78.0	ug/L	3	Standard
>	**Bi**	209	518292.4	0.9				ug/L	523984	Standard

Sample ID: QC Std 2
Report Date/Time: Wednesday, June 29, 2016 11:44:03
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	35.0	42.9	**-0.0179**	0.019	107.8	mg/L	37	Standard
K	39	36.7	39.4	**-0.0284**	0.035	124.9	mg/L	38	Standard
Ca	43	25.0	40.0	**-0.8569**	0.683	79.6	mg/L	10	Standard
Fe	54	125.6	8.4	**-0.0262**	0.013	49.1	mg/L	179	Standard
Fe	57	296.7	4.9	**-0.0150**	0.061	408.2	mg/L	335	Standard
Sc-1	45	58138.1	3.2				mg/L	61029	Standard
Cl	35	20966.5	2.9				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	1540.1	1.7				ug/L	1547	Standard
P	31	17027.9	2.4				ug/L	15286	Standard
S	34	1976.8	0.6				ug/L	1572	Standard
Sr	88	85.0	5.9				ug/L	73	Standard
C	12	30.0	33.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	6.3	86.9				mg/L	6	Standard
Ho-1	165	10.0	100.0				mg/L	10	Standard
Er	166	6.7	173.2				mg/L	13	Standard
I	127	1458.4	25.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.091	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 2
Report Date/Time: Wednesday, June 29, 2016 11:44:03
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.140
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	98.914
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 2	Sb	123	
QC Std 2	Ca	43	

Sample ID: QC Std 2
Report Date/Time: Wednesday, June 29, 2016 11:44:03
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 3
Sample Date/Time: Wednesday, June 29, 2016 11:46:53
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	66549.2	2.6				ug/L	74699	Standard
	Be	9	181.7	3.2	**0.2393**	0.009	3.6	ug/L	8	Standard
	Al	27	1430.1	59.0	**-0.0024**	0.007	267.2	ug/L	1493	Standard
	Sc	45	54693.6	2.6				ug/L	61029	Standard
	Ti	47	40.3	11.5	**0.0025**	0.013	506.9	ug/L	35	Standard
	V	51	4161.2	3.6	**0.3819**	0.016	4.1	ug/L	1435	Standard
	Cr	52	12463.1	1.9	**0.7390**	0.018	2.5	ug/L	7635	Standard
	Cr	53	996.7	1.0	**0.8006**	0.013	1.6	ug/L	363	Standard
	Mn	55	4091.6	2.3	**0.4679**	0.008	1.8	ug/L	747	Standard
	Co	59	2840.6	1.4	**0.4137**	0.010	2.4	ug/L	234	Standard
	Ni	60	3014.0	1.2	**1.5638**	0.035	2.2	ug/L	215	Standard
	Cu	65	1546.7	1.4	**0.7676**	0.020	2.7	ug/L	104	Standard
	Zn	66	7394.5	0.3	**6.6420**	0.044	0.7	ug/L	363	Standard
>	**Ge**	72	472336.8	0.9				ug/L	495142	Standard
	As	75	394.7	6.3	**0.4290**	0.024	5.7	ug/L	-50	Standard
	Se	82	54.4	27.0	**0.3818**	0.143	37.5	ug/L	14	Standard
	Se-1	77	75.3	14.2	**0.4599**	0.162	35.3	ug/L	42	Standard
>	**Ga**	71	58.3	44.0				mg/L	42	Standard
	Rb	85	8.3	34.6				ug/L	17	Standard
	Y	89	402582.8	1.9				ug/L	417098	Standard
>	**Rh**	103	5.0	100.0				ug/L	13	Standard
	Mo	98	114.4	34.0	**0.0400**	0.016	39.3	ug/L	11	Standard
	Ag	107	2464.9	1.7	**0.3900**	0.008	2.1	ug/L	68	Standard
	Cd	111	479.8	5.6	**0.2365**	0.013	5.7	mg/L	2	Standard
	Cd	114	1217.0	2.5	**0.2507**	0.007	2.6	ug/L	12	Standard
>	**In**	115	551471.6	0.4				ug/L	587441	Standard
	Sn	118	548.3	23.3	**0.0053**	0.024	443.1	ug/L	497	Standard
	Sb	123	2017.6	6.1	**0.4562**	0.029	6.4	ug/L	78	Standard
	Ba	135	1439.7	3.7	**0.7333**	0.029	4.0	ug/L	24	Standard
	Ce	140	11.7	49.5				ug/L	25	Standard
>	**Tb**	159	887425.9	1.8				ug/L	939701	Standard
	Ho	165	10.0	50.0				ug/L	10	Standard
	Tl	203	701.7	6.6	**0.0871**	0.005	5.3	ug/L	4	Standard
	Tl	205	588.3	14.1	**0.0861**	0.013	15.2	ug/L	2	Standard
	Pb	206	1333.4	2.9	**0.1894**	0.006	3.1	ug/L	240	Standard
	Pb	207	1083.0	1.4	**0.1683**	0.005	3.1	ug/L	198	Standard
	Pb	208	4575.3	2.2	**0.1806**	0.004	2.0	ug/L	883	Standard
	U	238	2677.9	3.2	**0.4017**	0.012	2.9	ug/L	3	Standard
>	**Bi**	209	501924.0	1.5				ug/L	523984	Standard

Sample ID: QC Std 3
Report Date/Time: Wednesday, June 29, 2016 11:49:09
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**21.9523**	38.014	173.2	mg/L	0	Standard
Mg	24	43.3	29.0	**-0.0023**	0.020	841.8	mg/L	37	Standard
K	39	33.3	37.7	**-0.0317**	0.032	101.1	mg/L	38	Standard
Ca	43	46.7	12.4	**0.7738**	0.482	62.2	mg/L	10	Standard
Fe	54	100.7	18.6	**-0.0432**	0.016	36.6	mg/L	179	Standard
Fe	57	325.0	8.1	**0.1578**	0.085	53.5	mg/L	335	Standard
> **Sc-1**	45	54693.6	2.6				mg/L	61029	Standard
Cl	35	21199.5	1.6				ug/L	20195	Standard
Kr	83	3.3	17.3				ug/L	4	Standard
Br	81	1403.4	5.3				ug/L	1547	Standard
P	31	11833.0	1.6				ug/L	15286	Standard
S	34	2000.1	8.0				ug/L	1572	Standard
Sr	88	93.3	20.3				ug/L	73	Standard
C	12	56.7	10.2				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	13.0	89.8				mg/L	6	Standard
Ho-1	165	10.0	50.0				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1430.1	34.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	119.656		
Al	27	-0.244		
Sc	45			
Ti	47			
V	51	95.473		
Cr	52	92.381		
Cr	53			
Mn	55	93.583		
Co	59	103.429		
Ni	60	97.737		
Cu	65	95.951		
Zn	66	106.273		
> Ge	72		95.394	
As	75	107.259		
Se	82	95.456		
Se-1	77			
> Ga	71			

Sample ID: QC Std 3
Report Date/Time: Wednesday, June 29, 2016 11:49:09
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	97.496		
Cd	111	98.556		
Cd	114			
> In	115			93.877
Sn	118			
Sb	123	114.045		
Ba	135	97.779		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	108.827		
Tl	205			
Pb	206			
Pb	207			
Pb	208	90.316		
U	238	100.429		
> Bi	209			95.790
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 3	Al	27	

Sample ID: QC Std 3
Report Date/Time: Wednesday, June 29, 2016 11:49:09
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Wednesday, June 29, 2016 11:50:53
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	71628.9	1.8				ug/L	74699	Standard
	Be	9	13.3	43.3	**0.0235**	0.007	30.3	ug/L	8	Standard
	Al	27	7079214.4	1.4	**50.2405**	1.571	3.1	ug/L	1493	Standard
	Sc	45	57092.4	1.6				ug/L	61029	Standard
	Ti	47	41841.7	1.2	**112.2257**	1.296	1.2	ug/L	35	Standard
	V	51	1975.7	11.7	**0.0706**	0.031	44.3	ug/L	1435	Standard
	Cr	52	7951.4	1.4	**0.0680**	0.015	21.7	ug/L	7635	Standard
	Cr	53	3292.0	3.5	**3.2946**	0.126	3.8	ug/L	363	Standard
	Mn	55	1443.4	4.1	**0.0629**	0.009	13.6	ug/L	747	Standard
	Co	59	431.3	5.1	**0.0292**	0.003	11.6	ug/L	234	Standard
	Ni	60	837.4	1.6	**0.3312**	0.007	2.2	ug/L	215	Standard
	Cu	65	638.0	4.2	**0.2451**	0.015	6.0	ug/L	104	Standard
	Zn	66	1716.8	2.4	**1.1448**	0.039	3.4	ug/L	363	Standard
>	Ge	72	486594.7	0.0				ug/L	495142	Standard
	As	75	-47.8	111.3	**0.0161**	0.048	300.1	ug/L	-50	Standard
	Se	82	12.5	17.3	**-0.0221**	0.020	90.2	ug/L	14	Standard
	Se-1	77	285.3	3.8	**3.3530**	0.150	4.5	ug/L	42	Standard
>	Ga	71	146.7	5.2				mg/L	42	Standard
	Rb	85	1056.7	1.8				ug/L	17	Standard
	Y	89	409909.2	0.2				ug/L	417098	Standard
>	Rh	103	11.7	99.0				ug/L	13	Standard
	Mo	98	258876.2	1.4	**101.0110**	2.466	2.4	ug/L	11	Standard
	Ag	107	250.7	8.2	**0.0271**	0.004	13.5	ug/L	68	Standard
	Cd	111	-191.4	6.3	**-0.1031**	0.005	5.1	mg/L	2	Standard
	Cd	114	476.2	9.0	**0.0945**	0.009	9.4	ug/L	12	Standard
>	In	115	564912.9	1.1				ug/L	587441	Standard
	Sn	118	765.0	2.0	**0.0415**	0.002	5.8	ug/L	497	Standard
	Sb	123	314.8	31.7	**0.0658**	0.022	32.8	ug/L	78	Standard
	Ba	135	628.0	0.0	**0.3031**	0.004	1.2	ug/L	24	Standard
	Ce	140	243.3	10.5				ug/L	25	Standard
>	Tb	159	934364.8	1.3				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	234.7	3.6	**0.0271**	0.001	3.8	ug/L	4	Standard
	Tl	205	175.0	22.3	**0.0250**	0.006	22.8	ug/L	2	Standard
	Pb	206	514.7	3.8	**0.0265**	0.004	16.6	ug/L	240	Standard
	Pb	207	427.7	5.6	**0.0214**	0.005	22.5	ug/L	198	Standard
	Pb	208	1691.4	2.8	**0.0174**	0.002	11.4	ug/L	883	Standard
	U	238	20.7	10.1	**-0.0014**	0.000	22.7	ug/L	3	Standard
>	Bi	209	516629.8	0.9				ug/L	523984	Standard

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 11:53:10
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**102.8074**	36.030	35.0	mg/L	0	Standard
Mg	24	9604.7	1.9	**13.1909**	0.366	2.8	mg/L	37	Standard
K	39	2033.5	6.6	**4.6893**	0.286	6.1	mg/L	38	Standard
Ca	43	250.0	4.0	**14.3615**	0.772	5.4	mg/L	10	Standard
Fe	54	14441.1	1.7	**12.9538**	0.416	3.2	mg/L	179	Standard
Fe	57	3845.5	3.6	**12.9047**	0.714	5.5	mg/L	335	Standard
> **Sc-1**	45	57092.4	1.6				mg/L	61029	Standard
Cl	35	20428.4	2.1				ug/L	20195	Standard
Kr	83	3.0	57.7				ug/L	4	Standard
Br	81	1493.4	8.7				ug/L	1547	Standard
P	31	15903.3	2.0				ug/L	15286	Standard
S	34	1818.4	8.5				ug/L	1572	Standard
Sr	88	70.0	35.7				ug/L	73	Standard
C	12	136.7	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	46.7	86.6				mg/L	3	Standard
Dy	164	12.2	93.4				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	23.3	24.7				mg/L	13	Standard
I	127	2226.8	19.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27	1.005		
Sc	45			
Ti	47	112.226		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.274	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 11:53:10
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	101.011		
Ag	107			
Cd	111			
Cd	114			
> In	115		96.165	
Sn	118			
Sb	123			
Ba	135			
Ce	140			
> Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
> Bi	209		98.596	
Na	23	822.459		
Mg	24	263.819		
K	39	93.787		
Ca	43	95.743		
Fe	54	103.630		
Fe	57	103.238		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Na	23	
QC Std 4	Mg	24	

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 11:53:10
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Wednesday, June 29, 2016 11:54:47
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72764.4	2.4				ug/L	74699	Standard
	Be	9	86093.9	0.6	**100.4334**	2.352	2.3	ug/L	8	Standard
	Al	27	7577089.0	1.5	**52.9410**	1.751	3.3	ug/L	1493	Standard
	Sc	45	54926.1	3.1				ug/L	61029	Standard
	Ti	47	43156.4	1.1	**115.6496**	0.651	0.6	ug/L	35	Standard
	V	51	767037.5	0.7	**103.0563**	0.476	0.5	ug/L	1435	Standard
	Cr	52	743370.8	0.7	**100.9527**	0.091	0.1	ug/L	7635	Standard
	Cr	53	95736.5	0.4	**104.9834**	0.488	0.5	ug/L	363	Standard
	Mn	55	719673.0	1.2	**104.9487**	1.961	1.9	ug/L	747	Standard
	Co	59	684381.8	1.6	**105.3392**	2.464	2.3	ug/L	234	Standard
	Ni	60	184775.9	0.8	**100.2112**	0.878	0.9	ug/L	215	Standard
	Cu	65	183945.6	0.7	**100.3790**	1.285	1.3	ug/L	104	Standard
	Zn	66	111716.4	0.5	**103.5231**	0.465	0.4	ug/L	363	Standard
>	**Ge**	72	487033.1	0.8				ug/L	495142	Standard
	As	75	115123.8	0.6	**104.5507**	0.234	0.2	ug/L	-50	Standard
	Se	82	11162.3	1.3	**103.2681**	1.937	1.9	ug/L	14	Standard
	Se-1	77	7790.3	1.4	**107.8133**	0.940	0.9	ug/L	42	Standard
>	**Ga**	71	155.0	24.4				mg/L	42	Standard
	Rb	85	853.4	12.0				ug/L	17	Standard
	Y	89	414937.1	1.1				ug/L	417098	Standard
>	**Rh**	103	48.3	21.5				ug/L	13	Standard
	Mo	98	271334.5	0.2	**106.0489**	0.711	0.7	ug/L	11	Standard
	Ag	107	596492.9	1.3	**95.3413**	1.752	1.8	ug/L	68	Standard
	Cd	111	202245.3	1.2	**100.7713**	1.934	1.9	mg/L	2	Standard
	Cd	114	489634.1	0.4	**99.4602**	0.649	0.7	ug/L	12	Standard
>	**In**	115	563893.1	0.8				ug/L	587441	Standard
	Sn	118	898.4	12.5	**0.0654**	0.019	28.7	ug/L	497	Standard
	Sb	123	448934.9	1.4	**100.1777**	2.039	2.0	ug/L	78	Standard
	Ba	135	193320.5	0.8	**98.3885**	1.494	1.5	ug/L	24	Standard
	Ce	140	68.3	16.9				ug/L	25	Standard
>	**Tb**	159	922182.5	1.6				ug/L	939701	Standard
	Ho	165	15.0	57.7				ug/L	10	Standard
	Tl	203	808287.0	0.7	**100.1763**	1.478	1.5	ug/L	4	Standard
	Tl	205	710155.2	1.4	**101.4182**	2.560	2.5	ug/L	2	Standard
	Pb	206	542765.1	0.5	**103.7300**	0.986	1.0	ug/L	240	Standard
	Pb	207	466015.8	1.0	**100.2432**	1.559	1.6	ug/L	198	Standard
	Pb	208	1879754.6	0.3	**102.3429**	1.262	1.2	ug/L	883	Standard
	U	238	673426.3	1.1	**99.9548**	2.202	2.2	ug/L	3	Standard
>	**Bi**	209	512933.6	1.2				ug/L	523984	Standard

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 11:57:04
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	10138.4	1.8	**14.4874**	0.596	4.1	mg/L	37	Standard
K	39	2125.1	3.3	**5.1060**	0.147	2.9	mg/L	38	Standard
Ca	43	265.0	8.2	**16.1201**	2.142	13.3	mg/L	10	Standard
Fe	54	14876.0	0.8	**13.8828**	0.353	2.5	mg/L	179	Standard
Fe	57	3958.9	4.0	**13.8986**	1.069	7.7	mg/L	335	Standard
Sc-1	45	54926.1	3.1				mg/L	61029	Standard
Cl	35	20351.0	1.2				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	1600.1	9.2				ug/L	1547	Standard
P	31	14895.7	0.4				ug/L	15286	Standard
S	34	1618.4	8.3				ug/L	1572	Standard
Sr	88	85.0	10.2				ug/L	73	Standard
C	12	116.7	71.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	16.2	38.3				mg/L	6	Standard
Ho-1	165	15.0	57.7				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	1496.7	25.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	100.433		
Al	27	1.059		
Sc	45			
Ti	47	115.650		
V	51	103.056		
Cr	52	100.953		
Cr	53			
Mn	55	104.949		
Co	59	105.339		
Ni	60	100.211		
Cu	65	100.379		
Zn	66	103.523		
Ge	72		98.362	
As	75	104.551		
Se	82	103.268		
Se-1	77			
Ga	71			

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 11:57:04
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	106.049		
Ag	107	95.341		
Cd	111	100.771		
Cd	114			
> In	115		95.991	
Sn	118			
Sb	123	100.178		
Ba	135	98.389		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	100.176		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.343		
U	238	99.955		
> Bi	209		97.891	
Na	23	0.040		
Mg	24	289.748		
K	39	102.121		
Ca	43	107.468		
Fe	54	111.063		
Fe	57	111.189		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Na	23	
QC Std 5	Mg	24	

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 11:57:04
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 12:00:10
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	74127.8	4.1				ug/L	74699	Standard
	Be	9	44057.3	2.0	**50.4723**	1.532	3.0	ug/L	8	Standard
	Al	27	7182862.5	0.8	**49.2846**	1.691	3.4	ug/L	1493	Standard
	Sc	45	54972.9	1.9				ug/L	61029	Standard
	Ti	47	37035.5	1.0	**100.2235**	3.375	3.4	ug/L	35	Standard
	V	51	380987.7	1.2	**51.6037**	1.920	3.7	ug/L	1435	Standard
	Cr	52	371095.9	0.5	**50.3905**	1.647	3.3	ug/L	7635	Standard
	Cr	53	45648.7	0.9	**50.3715**	0.927	1.8	ug/L	363	Standard
	Mn	55	362758.6	1.1	**53.3556**	2.033	3.8	ug/L	747	Standard
	Co	59	346170.1	1.6	**53.7955**	2.203	4.1	ug/L	234	Standard
	Ni	60	91822.1	1.5	**50.2407**	2.119	4.2	ug/L	215	Standard
	Cu	65	92043.5	2.0	**50.6873**	2.359	4.7	ug/L	104	Standard
	Zn	66	54708.9	1.1	**50.9762**	1.937	3.8	ug/L	363	Standard
>	Ge	72	482505.9	2.7				ug/L	495142	Standard
	As	75	56019.2	1.2	**51.4117**	1.775	3.5	ug/L	-50	Standard
	Se	82	5615.5	1.4	**52.4035**	2.100	4.0	ug/L	14	Standard
	Se-1	77	3787.5	1.5	**52.6293**	2.112	4.0	ug/L	42	Standard
>	Ga	71	63.3	9.1				mg/L	42	Standard
	Rb	85	771.7	5.9				ug/L	17	Standard
	Y	89	419939.4	2.7				ug/L	417098	Standard
>	Rh	103	31.7	9.1				ug/L	13	Standard
	Mo	98	262674.3	2.0	**102.5093**	6.034	5.9	ug/L	11	Standard
	Ag	107	318557.4	0.8	**50.8058**	2.007	4.0	ug/L	68	Standard
	Cd	111	103808.6	0.6	**51.6156**	2.157	4.2	mg/L	2	Standard
	Cd	114	253134.5	2.1	**51.2889**	1.034	2.0	ug/L	12	Standard
>	In	115	565595.5	4.0				ug/L	587441	Standard
	Sn	118	290084.1	1.0	**51.5381**	2.174	4.2	ug/L	497	Standard
	Sb	123	228198.9	1.3	**50.8307**	2.651	5.2	ug/L	78	Standard
	Ba	135	99366.1	0.6	**50.4579**	1.943	3.9	ug/L	24	Standard
	Ce	140	78.3	13.3				ug/L	25	Standard
>	Tb	159	913248.6	2.6				ug/L	939701	Standard
	Ho	165	3.3	86.6				ug/L	10	Standard
	Tl	203	413066.5	1.0	**51.9084**	1.016	2.0	ug/L	4	Standard
	Tl	205	363184.8	0.9	**52.5982**	1.678	3.2	ug/L	2	Standard
	Pb	206	268800.2	0.8	**52.0649**	1.553	3.0	ug/L	240	Standard
	Pb	207	239468.7	1.2	**52.2107**	1.790	3.4	ug/L	198	Standard
	Pb	208	958802.6	0.9	**52.9068**	1.662	3.1	ug/L	883	Standard
	U	238	349772.5	0.2	**52.6413**	1.317	2.5	ug/L	3	Standard
>	Bi	209	505954.2	2.3				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:02:27
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	3805.5	5.3	**5.3881**	0.260	4.8	mg/L	37	Standard
K	39	2106.8	1.9	**5.0574**	0.179	3.5	mg/L	38	Standard
Ca	43	101.7	5.7	**4.6110**	0.501	10.9	mg/L	10	Standard
Fe	54	5879.5	1.7	**5.3975**	0.172	3.2	mg/L	179	Standard
Fe	57	1785.1	7.3	**5.6694**	0.625	11.0	mg/L	335	Standard
> **Sc-1**	45	54972.9	1.9				mg/L	61029	Standard
Cl	35	21821.7	0.9				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	1470.1	4.5				ug/L	1547	Standard
P	31	16677.5	6.0				ug/L	15286	Standard
S	34	1955.1	4.5				ug/L	1572	Standard
Sr	88	73.3	14.2				ug/L	73	Standard
C	12	80.0	66.1				mg/L	63	Standard
N	14	6.7	86.6				mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.9	32.3				mg/L	6	Standard
Ho-1	165	3.3	86.6				mg/L	10	Standard
Er	166	16.7	124.9				mg/L	13	Standard
I	127	1668.4	28.6				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	100.945		
	Al	27	98.569		
	Sc	45			
	Ti	47	100.224		
	V	51	103.207		
	Cr	52	100.781		
	Cr	53			
	Mn	55	106.711		
	Co	59	107.591		
	Ni	60	100.481		
	Cu	65	101.375		
	Zn	66	101.952		
>	Ge	72		97.448	
	As	75	102.823		
	Se	82	104.807		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:02:27
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	102.509		
Ag	107	101.612		
Cd	111	103.231		
Cd	114			
> In	115			96.281
Sn	118	103.076		
Sb	123	101.661		
Ba	135	100.916		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	103.817		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.814		
U	238	105.283		
> Bi	209			96.559
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:02:27
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 12:07:34
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74348.6	1.0				ug/L	74699	Standard
	Be	9	21.7	26.6	**0.0324**	0.007	20.7	ug/L	8	Standard
	Al	27	1840.1	21.7	**-0.0008**	0.003	358.1	ug/L	1493	Standard
	Sc	45	56396.5	2.8				ug/L	61029	Standard
	Ti	47	36.7	15.0	**-0.0098**	0.014	147.0	ug/L	35	Standard
	V	51	1326.6	5.0	**-0.0153**	0.008	51.0	ug/L	1435	Standard
	Cr	52	6702.5	2.0	**-0.0955**	0.015	16.1	ug/L	7635	Standard
	Cr	53	393.3	15.3	**0.1070**	0.067	63.0	ug/L	363	Standard
	Mn	55	742.7	5.5	**-0.0386**	0.005	13.8	ug/L	747	Standard
	Co	59	250.3	1.5	**0.0016**	0.000	18.8	ug/L	234	Standard
	Ni	60	195.0	9.2	**-0.0170**	0.010	58.7	ug/L	215	Standard
	Cu	65	108.7	1.4	**-0.0438**	0.000	1.1	ug/L	104	Standard
	Zn	66	377.3	7.3	**-0.1000**	0.024	24.2	ug/L	363	Standard
>	**Ge**	72	482419.6	0.7				ug/L	495142	Standard
	As	75	-16.3	122.1	**0.0447**	0.018	40.8	ug/L	-50	Standard
	Se	82	12.2	10.1	**-0.0240**	0.012	49.4	ug/L	14	Standard
	Se-1	77	52.7	12.2	**0.1177**	0.089	75.6	ug/L	42	Standard
>	**Ga**	71	61.7	33.8				mg/L	42	Standard
	Rb	85	30.0	50.0				ug/L	17	Standard
	Y	89	414790.7	2.9				ug/L	417098	Standard
>	**Rh**	103	1.7	173.2				ug/L	13	Standard
	Mo	98	135.9	9.5	**0.0469**	0.005	10.2	ug/L	11	Standard
	Ag	107	96.3	7.3	**0.0024**	0.001	45.2	ug/L	68	Standard
	Cd	111	7.4	43.0	**-0.0042**	0.002	37.1	mg/L	2	Standard
	Cd	114	24.8	51.6	**0.0029**	0.003	88.9	ug/L	12	Standard
>	**In**	115	568967.0	0.5				ug/L	587441	Standard
	Sn	118	628.3	16.1	**0.0164**	0.018	110.8	ug/L	497	Standard
	Sb	123	617.2	40.6	**0.1324**	0.056	42.3	ug/L	78	Standard
	Ba	135	30.7	16.1	**-0.0005**	0.003	469.7	ug/L	24	Standard
	Ce	140	23.3	68.9				ug/L	25	Standard
>	**Tb**	159	918183.5	1.2				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	30.3	14.9	**0.0020**	0.001	31.1	ug/L	4	Standard
	Tl	205	23.3	24.7	**0.0035**	0.001	22.2	ug/L	2	Standard
	Pb	206	259.3	0.6	**-0.0221**	0.001	4.2	ug/L	240	Standard
	Pb	207	198.7	1.6	**-0.0277**	0.001	4.0	ug/L	198	Standard
	Pb	208	894.0	2.8	**-0.0259**	0.002	7.7	ug/L	883	Standard
	U	238	15.0	35.3	**-0.0022**	0.001	35.0	ug/L	3	Standard
>	**Bi**	209	518580.7	1.8				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:09:51
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**20.6253**	35.715	173.2	mg/L	0	Standard
Mg	24	31.7	32.9	**-0.0207**	0.015	73.7	mg/L	37	Standard
K	39	51.7	24.4	**0.0090**	0.027	299.1	mg/L	38	Standard
Ca	43	33.3	31.2	**-0.2319**	0.781	336.6	mg/L	10	Standard
Fe	54	140.8	7.6	**-0.0093**	0.006	67.4	mg/L	179	Standard
Fe	57	256.7	18.9	**-0.1295**	0.180	138.6	mg/L	335	Standard
Sc-1	45	56396.5	2.8				mg/L	61029	Standard
Cl	35	21372.4	1.3				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	1540.1	8.8				ug/L	1547	Standard
P	31	15931.7	0.9				ug/L	15286	Standard
S	34	1863.4	1.4				ug/L	1572	Standard
Sr	88	58.3	17.8				ug/L	73	Standard
C	12	56.7	27.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	6.2	186.6				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	10.0					mg/L	13	Standard
I	127	1518.4	26.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		97.431	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:09:51
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.855
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.969
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:09:51
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066106
Sample Date/Time: Wednesday, June 29, 2016 12:12:15
Number of Replicates: 3
Autosampler Position: 301
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78301.9	3.0				ug/L	74699	Standard
	Be	9	591.7	15.8	**0.6482**	0.094	14.5	ug/L	8	Standard
	Al	27	14033022.6	1.2	**91.1117**	1.566	1.7	ug/L	1493	Standard
	Sc	45	64031.7	1.2				ug/L	61029	Standard
	Ti	47	6936.9	1.2	**18.7002**	0.397	2.1	ug/L	35	Standard
	V	51	136379.5	0.6	**18.3618**	0.342	1.9	ug/L	1435	Standard
	Cr	52	135090.6	0.2	**17.7119**	0.504	2.8	ug/L	7635	Standard
	Cr	53	16100.2	3.4	**17.5840**	1.043	5.9	ug/L	363	Standard
	Mn	55	2727656.7	0.9	**402.5207**	10.967	2.7	ug/L	747	Standard
	Co	59	46530.0	0.2	**7.2050**	0.195	2.7	ug/L	234	Standard
	Ni	60	31243.2	1.0	**17.0295**	0.585	3.4	ug/L	215	Standard
	Cu	65	40278.5	0.5	**22.1396**	0.604	2.7	ug/L	104	Standard
	Zn	66	62615.1	0.3	**58.4626**	1.558	2.7	ug/L	363	Standard
>	**Ge**	72	481946.2	2.4				ug/L	495142	Standard
	As	75	4644.7	1.4	**4.3222**	0.154	3.6	ug/L	-50	Standard
	Se	82	54.0	14.6	**0.3688**	0.086	23.4	ug/L	14	Standard
	Se-1	77	96.0	5.2	**0.7275**	0.040	5.5	ug/L	42	Standard
>	**Ga**	71	12810.4	4.6				mg/L	42	Standard
	Rb	85	116256.0	0.7				ug/L	17	Standard
	Y	89	585719.0	1.0				ug/L	417098	Standard
>	**Rh**	103	45.0	33.3				ug/L	13	Standard
	Mo	98	935.2	3.6	**0.3550**	0.010	2.8	ug/L	11	Standard
	Ag	107	931.4	1.2	**0.1340**	0.001	0.5	ug/L	68	Standard
	Cd	111	782.7	0.6	**0.3771**	0.007	1.8	mg/L	2	Standard
	Cd	114	1959.7	7.4	**0.3906**	0.025	6.5	ug/L	12	Standard
>	**In**	115	571359.2	1.2				ug/L	587441	Standard
	Sn	118	1011.7	7.7	**0.0833**	0.012	14.7	ug/L	497	Standard
	Sb	123	336.5	28.9	**0.0698**	0.021	29.4	ug/L	78	Standard
	Ba	135	394959.7	1.6	**198.3814**	1.476	0.7	ug/L	24	Standard
	Ce	140	893729.8	1.7				ug/L	25	Standard
>	**Tb**	159	948738.0	1.3				ug/L	939701	Standard
	Ho	165	15724.8	2.6				ug/L	10	Standard
	Tl	203	658.0	2.7	**0.0791**	0.002	2.2	ug/L	4	Standard
	Tl	205	583.3	11.6	**0.0828**	0.010	12.2	ug/L	2	Standard
	Pb	206	370590.1	1.2	**70.2116**	0.816	1.2	ug/L	240	Standard
	Pb	207	304396.4	0.8	**64.9093**	0.988	1.5	ug/L	198	Standard
	Pb	208	1254498.7	1.7	**67.7133**	1.716	2.5	ug/L	883	Standard
	U	238	3382.4	2.2	**0.4935**	0.016	3.3	ug/L	3	Standard
>	**Bi**	209	517229.2	1.1				ug/L	523984	Standard

Sample ID: L1606066106
Report Date/Time: Wednesday, June 29, 2016 12:14:32
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**36.6019**	63.388	173.2	mg/L	0	Standard
Mg	24	128.3	15.7	**0.0927**	0.024	25.5	mg/L	37	Standard
K	39	566.7	14.3	**1.0807**	0.180	16.7	mg/L	38	Standard
Ca	43	418.3	12.7	**22.6893**	3.491	15.4	mg/L	10	Standard
Fe	54	12793.3	3.0	**10.1983**	0.198	1.9	mg/L	179	Standard
Fe	57	3678.8	1.9	**10.8461**	0.363	3.4	mg/L	335	Standard
Sc-1	45	64031.7	1.2				mg/L	61029	Standard
Cl	35	22050.7	1.1				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2173.5	11.7				ug/L	1547	Standard
P	31	16252.1	5.4				ug/L	15286	Standard
S	34	2065.1	4.4				ug/L	1572	Standard
Sr	88	140.0	9.4				ug/L	73	Standard
C	12	143.3	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	203.3	10.2				mg/L	3	Standard
Dy	164	25060.4	3.8				mg/L	6	Standard
Ho-1	165	15724.8	2.6				mg/L	10	Standard
Er	166	13529.4	4.1				mg/L	13	Standard
I	127	8297.3	2.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		104.824	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		97.335	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606066106
Report Date/Time: Wednesday, June 29, 2016 12:14:32
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.262
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	98.711
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606066106
Report Date/Time: Wednesday, June 29, 2016 12:14:32
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066107
Sample Date/Time: Wednesday, June 29, 2016 12:15:26
Number of Replicates: 3
Autosampler Position: 302
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79855.0	3.3				ug/L	74699	Standard
	Be	9	765.0	6.8	**0.8219**	0.077	9.4	ug/L	8	Standard
	Al	27	26230756.6	0.7	**167.0452**	4.333	2.6	ug/L	1493	Standard
	Sc	45	64585.8	4.1				ug/L	61029	Standard
	Ti	47	7966.1	1.8	**21.8574**	0.352	1.6	ug/L	35	Standard
	V	51	155637.7	1.4	**21.3440**	0.459	2.1	ug/L	1435	Standard
	Cr	52	137713.7	0.8	**18.3989**	0.313	1.7	ug/L	7635	Standard
	Cr	53	16806.0	3.9	**18.6780**	0.841	4.5	ug/L	363	Standard
	Mn	55	4726467.2	1.1	**709.3402**	2.876	0.4	ug/L	747	Standard
	Co	59	64550.2	0.1	**10.1794**	0.081	0.8	ug/L	234	Standard
	Ni	60	42709.8	0.1	**23.7183**	0.226	1.0	ug/L	215	Standard
	Cu	65	23683.8	1.0	**13.1970**	0.251	1.9	ug/L	104	Standard
	Zn	66	44795.8	1.0	**42.4101**	0.817	1.9	ug/L	363	Standard
>	**Ge**	72	473762.9	0.9				ug/L	495142	Standard
	As	75	7180.0	1.6	**6.7588**	0.066	1.0	ug/L	-50	Standard
	Se	82	142.2	5.2	**1.2157**	0.067	5.5	ug/L	14	Standard
	Se-1	77	110.3	7.6	**0.9569**	0.126	13.2	ug/L	42	Standard
>	**Ga**	71	17858.9	0.4				mg/L	42	Standard
	Rb	85	152815.2	1.7				ug/L	17	Standard
	Y	89	630725.5	1.5				ug/L	417098	Standard
>	**Rh**	103	41.7	25.0				ug/L	13	Standard
	Mo	98	1176.0	1.0	**0.4476**	0.010	2.2	ug/L	11	Standard
	Ag	107	471.0	1.7	**0.0613**	0.002	3.5	ug/L	68	Standard
	Cd	111	231.8	9.8	**0.1061**	0.013	12.1	mg/L	2	Standard
	Cd	114	548.2	12.6	**0.1078**	0.015	13.9	ug/L	12	Standard
>	**In**	115	571871.5	1.5				ug/L	587441	Standard
	Sn	118	1063.4	7.5	**0.0923**	0.015	16.1	ug/L	497	Standard
	Sb	123	198.5	41.2	**0.0396**	0.018	46.2	ug/L	78	Standard
	Ba	135	410562.2	0.2	**206.0643**	2.723	1.3	ug/L	24	Standard
	Ce	140	1020863.3	0.8				ug/L	25	Standard
>	**Tb**	159	935971.5	1.4				ug/L	939701	Standard
	Ho	165	18616.4	1.9				ug/L	10	Standard
	Tl	203	969.7	3.1	**0.1217**	0.006	4.6	ug/L	4	Standard
	Tl	205	781.7	5.1	**0.1148**	0.005	4.1	ug/L	2	Standard
	Pb	206	74583.8	0.8	**14.5751**	0.129	0.9	ug/L	240	Standard
	Pb	207	60293.4	1.5	**13.2552**	0.050	0.4	ug/L	198	Standard
	Pb	208	253434.0	0.7	**14.1059**	0.309	2.2	ug/L	883	Standard
	U	238	4763.7	0.8	**0.7216**	0.012	1.7	ug/L	3	Standard
>	**Bi**	209	499544.2	1.5				ug/L	523984	Standard

Sample ID: L1606066107
Report Date/Time: Wednesday, June 29, 2016 12:17:43
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**90.1824**	29.248	32.4	mg/L	0	Standard
Mg	24	323.3	10.9	**0.3294**	0.041	12.6	mg/L	37	Standard
K	39	788.4	10.3	**1.5298**	0.110	7.2	mg/L	38	Standard
Ca	43	663.3	6.3	**37.0621**	1.711	4.6	mg/L	10	Standard
Fe	54	18630.5	1.2	**14.8018**	0.561	3.8	mg/L	179	Standard
Fe	57	5110.9	3.8	**15.3724**	1.217	7.9	mg/L	335	Standard
Sc-1	45	64585.8	4.1				mg/L	61029	Standard
Cl	35	22235.6	2.9				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	23104.3	4.2				ug/L	1547	Standard
P	31	15829.9	2.1				ug/L	15286	Standard
S	34	2033.5	3.2				ug/L	1572	Standard
Sr	88	161.7	12.5				ug/L	73	Standard
C	12	96.7	41.8				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	31177.2	1.6				mg/L	6	Standard
Ho-1	165	18616.4	1.9				mg/L	10	Standard
Er	166	16976.2	1.7				mg/L	13	Standard
I	127	9242.8	4.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		106.903	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.682	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606066107
Report Date/Time: Wednesday, June 29, 2016 12:17:43
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	97.350	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	95.336	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606066107
Report Date/Time: Wednesday, June 29, 2016 12:17:43
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066108
Sample Date/Time: Wednesday, June 29, 2016 12:18:38
Number of Replicates: 3
Autosampler Position: 303
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	77986.9	1.5				ug/L	74699	Standard
	Be	9	848.4	8.3	**0.9313**	0.087	9.4	ug/L	8	Standard
	Al	27	17300686.4	1.1	**112.7510**	0.475	0.4	ug/L	1493	Standard
	Sc	45	64965.7	2.0				ug/L	61029	Standard
	Ti	47	6993.9	0.6	**18.3803**	0.198	1.1	ug/L	35	Standard
	V	51	170039.3	0.3	**22.3624**	0.095	0.4	ug/L	1435	Standard
	Cr	52	135852.7	1.1	**17.3429**	0.173	1.0	ug/L	7635	Standard
	Cr	53	16400.5	1.1	**17.4496**	0.150	0.9	ug/L	363	Standard
	Mn	55	3732645.2	0.8	**537.0232**	5.661	1.1	ug/L	747	Standard
	Co	59	59499.9	1.1	**8.9910**	0.146	1.6	ug/L	234	Standard
	Ni	60	38977.4	1.0	**20.7345**	0.185	0.9	ug/L	215	Standard
	Cu	65	32324.4	0.5	**17.2976**	0.098	0.6	ug/L	104	Standard
	Zn	66	55484.7	0.8	**50.4392**	0.445	0.9	ug/L	363	Standard
>	Ge	72	494184.7	0.5				ug/L	495142	Standard
	As	75	5935.1	0.6	**5.3688**	0.057	1.1	ug/L	-50	Standard
	Se	82	56.5	10.0	**0.3782**	0.052	13.7	ug/L	14	Standard
	Se-1	77	118.0	0.8	**0.9965**	0.019	1.9	ug/L	42	Standard
>	Ga	71	16687.5	1.9				mg/L	42	Standard
	Rb	85	157851.4	1.2				ug/L	17	Standard
	Y	89	635672.9	1.6				ug/L	417098	Standard
>	Rh	103	68.3	27.7				ug/L	13	Standard
	Mo	98	1214.1	0.7	**0.4777**	0.002	0.4	ug/L	11	Standard
	Ag	107	767.0	3.0	**0.1120**	0.004	3.7	ug/L	68	Standard
	Cd	111	562.6	3.0	**0.2776**	0.008	3.0	mg/L	2	Standard
	Cd	114	1264.6	12.4	**0.2596**	0.032	12.4	ug/L	12	Standard
>	In	115	553515.9	0.3				ug/L	587441	Standard
	Sn	118	763.4	11.0	**0.0440**	0.015	34.7	ug/L	497	Standard
	Sb	123	242.9	15.0	**0.0510**	0.008	16.4	ug/L	78	Standard
	Ba	135	429181.6	1.6	**222.5293**	3.912	1.8	ug/L	24	Standard
	Ce	140	1102655.1	2.3				ug/L	25	Standard
>	Tb	159	923824.1	1.4				ug/L	939701	Standard
	Ho	165	18663.2	2.9				ug/L	10	Standard
	Tl	203	782.7	3.6	**0.0981**	0.005	5.3	ug/L	4	Standard
	Tl	205	701.7	8.6	**0.1034**	0.010	10.1	ug/L	2	Standard
	Pb	206	213507.8	0.5	**41.9534**	0.898	2.1	ug/L	240	Standard
	Pb	207	175202.2	0.5	**38.7416**	0.683	1.8	ug/L	198	Standard
	Pb	208	727362.3	0.6	**40.7140**	1.033	2.5	ug/L	883	Standard
	U	238	4175.6	1.0	**0.6335**	0.018	2.9	ug/L	3	Standard
>	Bi	209	498479.4	1.9				ug/L	523984	Standard

Sample ID: L1606066108
Report Date/Time: Wednesday, June 29, 2016 12:20:55
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.0364**	31.231	173.2	mg/L	0	Standard
Mg	24	140.0	21.7	**0.1047**	0.037	35.2	mg/L	37	Standard
K	39	708.4	12.8	**1.3550**	0.168	12.4	mg/L	38	Standard
Ca	43	508.3	5.0	**27.6291**	0.895	3.2	mg/L	10	Standard
Fe	54	15324.4	1.1	**12.0707**	0.309	2.6	mg/L	179	Standard
Fe	57	4262.3	3.3	**12.5338**	0.279	2.2	mg/L	335	Standard
> **Sc-1**	45	64965.7	2.0				mg/L	61029	Standard
Cl	35	21927.9	1.6				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	2823.6	3.3				ug/L	1547	Standard
P	31	16612.4	1.0				ug/L	15286	Standard
S	34	2085.1	7.7				ug/L	1572	Standard
Sr	88	245.0	14.1				ug/L	73	Standard
C	12	116.7	27.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	93.3	49.5				mg/L	3	Standard
Dy	164	30954.0	1.2				mg/L	6	Standard
Ho-1	165	18663.2	2.9				mg/L	10	Standard
Er	166	16615.8	2.6				mg/L	13	Standard
I	127	11766.2	5.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		104.402	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		99.807	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606066108
Report Date/Time: Wednesday, June 29, 2016 12:20:55
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	94.225
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.132
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606066108
Report Date/Time: Wednesday, June 29, 2016 12:20:55
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066109
Sample Date/Time: Wednesday, June 29, 2016 12:21:49
Number of Replicates: 3
Autosampler Position: 304
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77037.0	1.5				ug/L	74699	Standard
	Be	9	378.3	6.2	**0.4242**	0.022	5.2	ug/L	8	Standard
	Al	27	13254869.5	1.3	**87.4672**	2.449	2.8	ug/L	1493	Standard
	Sc	45	63133.0	3.0				ug/L	61029	Standard
	Ti	47	8144.2	1.7	**21.2616**	0.854	4.0	ug/L	35	Standard
	V	51	109381.5	1.5	**14.2029**	0.301	2.1	ug/L	1435	Standard
	Cr	52	127693.7	2.3	**16.1132**	0.755	4.7	ug/L	7635	Standard
	Cr	53	15149.2	3.9	**15.9774**	1.000	6.3	ug/L	363	Standard
	Mn	55	3010101.1	0.9	**429.7771**	13.412	3.1	ug/L	747	Standard
	Co	59	32158.1	0.7	**4.8045**	0.098	2.0	ug/L	234	Standard
	Ni	60	23283.5	0.8	**12.2424**	0.383	3.1	ug/L	215	Standard
	Cu	65	9940.9	0.4	**5.2071**	0.131	2.5	ug/L	104	Standard
	Zn	66	24785.3	0.2	**22.1054**	0.470	2.1	ug/L	363	Standard
>	**Ge**	72	498160.0	2.3				ug/L	495142	Standard
	As	75	2943.4	0.7	**2.6727**	0.077	2.9	ug/L	-50	Standard
	Se	82	39.1	13.2	**0.2170**	0.054	24.7	ug/L	14	Standard
	Se-1	77	75.7	11.2	**0.4068**	0.103	25.3	ug/L	42	Standard
>	**Ga**	71	9034.4	2.0				mg/L	42	Standard
	Rb	85	68682.0	2.5				ug/L	17	Standard
	Y	89	539979.1	1.9				ug/L	417098	Standard
>	**Rh**	103	11.7	24.7				ug/L	13	Standard
	Mo	98	3058.1	1.5	**1.1809**	0.018	1.5	ug/L	11	Standard
	Ag	107	334.3	34.2	**0.0400**	0.017	43.3	ug/L	68	Standard
	Cd	111	137.9	39.4	**0.0601**	0.026	43.1	mg/L	2	Standard
	Cd	114	340.6	39.5	**0.0664**	0.026	39.4	ug/L	12	Standard
>	**In**	115	568023.6	1.3				ug/L	587441	Standard
	Sn	118	973.4	3.3	**0.0776**	0.003	4.4	ug/L	497	Standard
	Sb	123	217.2	33.7	**0.0438**	0.016	35.6	ug/L	78	Standard
	Ba	135	138215.3	1.3	**69.8220**	0.106	0.2	ug/L	24	Standard
	Ce	140	545871.5	1.0				ug/L	25	Standard
>	**Tb**	159	946980.1	0.3				ug/L	939701	Standard
	Ho	165	11547.7	1.8				ug/L	10	Standard
	Tl	203	458.7	30.8	**0.0539**	0.017	31.9	ug/L	4	Standard
	Tl	205	360.0	14.1	**0.0505**	0.007	14.1	ug/L	2	Standard
	Pb	206	30575.5	1.0	**5.6549**	0.063	1.1	ug/L	240	Standard
	Pb	207	24590.3	1.3	**5.1132**	0.071	1.4	ug/L	198	Standard
	Pb	208	101567.1	0.7	**5.3447**	0.038	0.7	ug/L	883	Standard
	U	238	5591.0	1.3	**0.8082**	0.011	1.4	ug/L	3	Standard
>	**Bi**	209	523773.4	0.1				ug/L	523984	Standard

Sample ID: L1606066109
Report Date/Time: Wednesday, June 29, 2016 12:24:06
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.1528**	33.165	173.2	mg/L	0	Standard
Mg	24	88.3	26.1	**0.0447**	0.026	57.5	mg/L	37	Standard
K	39	410.0	24.0	**0.7638**	0.217	28.4	mg/L	38	Standard
Ca	43	368.3	13.4	**20.0310**	3.643	18.2	mg/L	10	Standard
Fe	54	11096.4	1.9	**8.9581**	0.101	1.1	mg/L	179	Standard
Fe	57	2985.3	3.6	**8.7334**	0.110	1.3	mg/L	335	Standard
Sc-1	45	63133.0	3.0				mg/L	61029	Standard
Cl	35	21859.8	1.0				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	1696.8	5.1				ug/L	1547	Standard
P	31	15684.8	2.6				ug/L	15286	Standard
S	34	1883.5	4.2				ug/L	1572	Standard
Sr	88	116.7	19.3				ug/L	73	Standard
C	12	100.0	26.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	18768.7	3.5				mg/L	6	Standard
Ho-1	165	11547.7	1.8				mg/L	10	Standard
Er	166	10790.5	1.6				mg/L	13	Standard
I	127	6703.2	5.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		103.130	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		100.609	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606066109
Report Date/Time: Wednesday, June 29, 2016 12:24:06
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.695
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.960
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606066109
Report Date/Time: Wednesday, June 29, 2016 12:24:06
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082203
Sample Date/Time: Wednesday, June 29, 2016 12:25:01
Number of Replicates: 3
Autosampler Position: 305
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75246.5	2.8				ug/L	74699	Standard
	Be	9	6.7	114.6	**0.0150**	0.008	55.4	ug/L	8	Standard
	Al	27	1140453.8	1.3	**7.6935**	0.186	2.4	ug/L	1493	Standard
	Sc	45	58880.9	1.2				ug/L	61029	Standard
	Ti	47	77.7	5.2	**0.0982**	0.010	10.0	ug/L	35	Standard
	V	51	1502.7	7.0	**0.0058**	0.017	290.5	ug/L	1435	Standard
	Cr	52	10404.9	0.3	**0.3964**	0.023	5.7	ug/L	7635	Standard
	Cr	53	878.4	11.3	**0.6323**	0.120	19.0	ug/L	363	Standard
	Mn	55	64288.8	1.0	**9.1920**	0.174	1.9	ug/L	747	Standard
	Co	59	1249.1	3.1	**0.1540**	0.006	4.1	ug/L	234	Standard
	Ni	60	673.3	4.1	**0.2398**	0.011	4.6	ug/L	215	Standard
	Cu	65	229.7	8.0	**0.0211**	0.010	48.0	ug/L	104	Standard
	Zn	66	2146.2	1.5	**1.5328**	0.038	2.5	ug/L	363	Standard
>	**Ge**	72	489582.8	1.5				ug/L	495142	Standard
	As	75	99.3	9.7	**0.1492**	0.008	5.6	ug/L	-50	Standard
	Se	82	43.5	22.3	**0.2617**	0.085	32.3	ug/L	14	Standard
	Se-1	77	67.7	6.2	**0.3148**	0.059	18.8	ug/L	42	Standard
>	**Ga**	71	61.7	40.8				mg/L	42	Standard
	Rb	85	88606.7	2.0				ug/L	17	Standard
	Y	89	430024.7	1.6				ug/L	417098	Standard
>	**Rh**	103	28.3	50.9				ug/L	13	Standard
	Mo	98	81.5	16.2	**0.0249**	0.005	20.1	ug/L	11	Standard
	Ag	107	97.7	18.1	**0.0021**	0.003	124.3	ug/L	68	Standard
	Cd	111	13.9	57.3	**-0.0013**	0.004	291.1	mg/L	2	Standard
	Cd	114	23.0	35.9	**0.0024**	0.002	67.9	ug/L	12	Standard
>	**In**	115	586341.4	0.7				ug/L	587441	Standard
	Sn	118	508.3	13.9	**-0.0076**	0.011	151.7	ug/L	497	Standard
	Sb	123	156.7	24.2	**0.0294**	0.008	27.7	ug/L	78	Standard
	Ba	135	76489.9	1.2	**37.4252**	0.302	0.8	ug/L	24	Standard
	Ce	140	665.0	13.3				ug/L	25	Standard
>	**Tb**	159	934158.4	1.1				ug/L	939701	Standard
	Ho	165	15.0	66.7				ug/L	10	Standard
	Tl	203	113.7	11.3	**0.0121**	0.002	13.7	ug/L	4	Standard
	Tl	205	93.3	44.6	**0.0133**	0.006	44.3	ug/L	2	Standard
	Pb	206	362.0	5.5	**-0.0029**	0.004	131.1	ug/L	240	Standard
	Pb	207	281.3	6.6	**-0.0103**	0.004	38.7	ug/L	198	Standard
	Pb	208	1197.0	7.1	**-0.0098**	0.005	49.2	ug/L	883	Standard
	U	238	25.0	34.9	**-0.0008**	0.001	168.5	ug/L	3	Standard
>	**Bi**	209	520173.1	0.7				ug/L	523984	Standard

Sample ID: L1606082203
Report Date/Time: Wednesday, June 29, 2016 12:27:18
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	3547.1	5.1	**4.6793**	0.186	4.0	mg/L	37	Standard
K	39	1096.7	1.1	**2.3987**	0.055	2.3	mg/L	38	Standard
Ca	43	51.7	27.9	**0.8562**	0.931	108.7	mg/L	10	Standard
Fe	54	433.8	1.2	**0.2430**	0.009	3.5	mg/L	179	Standard
Fe	57	376.7	13.8	**0.2532**	0.189	74.5	mg/L	335	Standard
> **Sc-1**	45	58880.9	1.2				mg/L	61029	Standard
Cl	35	22906.6	0.4				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	9496.3	3.0				ug/L	1547	Standard
P	31	16589.1	2.9				ug/L	15286	Standard
S	34	2071.8	1.6				ug/L	1572	Standard
Sr	88	118.3	17.1				ug/L	73	Standard
C	12	53.3	10.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	19.0	47.5				mg/L	6	Standard
Ho-1	165	15.0	66.7				mg/L	10	Standard
Er	166	20.0	100.0				mg/L	13	Standard
I	127	14315.1	5.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		100.733	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.877	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082203
Report Date/Time: Wednesday, June 29, 2016 12:27:18
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.813
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.273
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082302
Sample Date/Time: Wednesday, June 29, 2016 12:28:12
Number of Replicates: 3
Autosampler Position: 306
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75794.1	0.6				ug/L	74699	Standard
	Be	9	13.3	21.7	**0.0225**	0.003	14.2	ug/L	8	Standard
	Al	27	2805598.0	0.7	**18.8025**	0.242	1.3	ug/L	1493	Standard
	Sc	45	59848.0	1.2				ug/L	61029	Standard
	Ti	47	42.3	20.1	**0.0017**	0.021	1272.3	ug/L	35	Standard
	V	51	1764.1	8.5	**0.0361**	0.018	49.9	ug/L	1435	Standard
	Cr	52	8282.9	3.5	**0.0853**	0.037	43.4	ug/L	7635	Standard
	Cr	53	906.7	14.2	**0.6437**	0.138	21.5	ug/L	363	Standard
	Mn	55	61274.7	1.5	**8.5848**	0.068	0.8	ug/L	747	Standard
	Co	59	3907.8	2.1	**0.5500**	0.015	2.7	ug/L	234	Standard
	Ni	60	654.7	3.9	**0.2231**	0.012	5.6	ug/L	215	Standard
	Cu	65	222.3	1.8	**0.0148**	0.001	8.1	ug/L	104	Standard
	Zn	66	2204.5	3.0	**1.5481**	0.054	3.5	ug/L	363	Standard
>	**Ge**	72	498979.3	0.8				ug/L	495142	Standard
	As	75	72.7	46.8	**0.1239**	0.030	24.0	ug/L	-50	Standard
	Se	82	52.0	6.1	**0.3317**	0.029	8.6	ug/L	14	Standard
	Se-1	77	105.0	8.3	**0.8037**	0.108	13.5	ug/L	42	Standard
>	**Ga**	71	56.7	13.5				mg/L	42	Standard
	Rb	85	343.3	5.1				ug/L	17	Standard
	Y	89	428520.0	1.9				ug/L	417098	Standard
>	**Rh**	103	38.3	27.2				ug/L	13	Standard
	Mo	98	38.4	23.3	**0.0089**	0.004	39.5	ug/L	11	Standard
	Ag	107	77.7	2.7	**-0.0008**	0.000	47.2	ug/L	68	Standard
	Cd	111	5.6	57.3	**-0.0052**	0.002	30.1	mg/L	2	Standard
	Cd	114	18.7	97.5	**0.0016**	0.004	227.0	ug/L	12	Standard
>	**In**	115	580343.3	0.7				ug/L	587441	Standard
	Sn	118	485.0	23.2	**-0.0106**	0.020	190.8	ug/L	497	Standard
	Sb	123	114.5	31.8	**0.0207**	0.008	39.1	ug/L	78	Standard
	Ba	135	150889.3	1.2	**74.6061**	0.512	0.7	ug/L	24	Standard
	Ce	140	123.3	15.3				ug/L	25	Standard
>	**Tb**	159	936316.1	1.2				ug/L	939701	Standard
	Ho	165	18.3	31.5				ug/L	10	Standard
	Tl	203	189.0	9.5	**0.0214**	0.002	8.8	ug/L	4	Standard
	Tl	205	161.7	6.4	**0.0230**	0.002	7.6	ug/L	2	Standard
	Pb	206	327.3	6.0	**-0.0093**	0.004	47.2	ug/L	240	Standard
	Pb	207	269.3	8.1	**-0.0127**	0.005	36.6	ug/L	198	Standard
	Pb	208	1102.3	4.7	**-0.0147**	0.003	21.5	ug/L	883	Standard
	U	238	62.7	10.4	**0.0048**	0.001	17.9	ug/L	3	Standard
>	**Bi**	209	519166.0	1.5				ug/L	523984	Standard

Sample ID: L1606082302
Report Date/Time: Wednesday, June 29, 2016 12:30:29
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	5325.9	0.2	**6.9463**	0.099	1.4	mg/L	37	Standard
K	39	75.0	37.1	**0.0544**	0.061	111.8	mg/L	38	Standard
Ca	43	86.7	3.3	**3.0583**	0.130	4.3	mg/L	10	Standard
Fe	54	170.8	11.8	**0.0094**	0.018	186.9	mg/L	179	Standard
Fe	57	311.7	10.3	**0.0068**	0.121	1775.6	mg/L	335	Standard
> **Sc-1**	45	59848.0	1.2				mg/L	61029	Standard
Cl	35	23747.3	1.3				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	11654.5	4.8				ug/L	1547	Standard
P	31	16051.8	4.0				ug/L	15286	Standard
S	34	1918.5	1.8				ug/L	1572	Standard
Sr	88	140.0	9.4				ug/L	73	Standard
C	12	53.3	39.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	12.2	186.8				mg/L	6	Standard
Ho-1	165	18.3	31.5				mg/L	10	Standard
Er	166	23.3	24.7				mg/L	13	Standard
I	127	13899.7	5.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.466	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.775	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082302
Report Date/Time: Wednesday, June 29, 2016 12:30:29
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.792
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.080
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082303
Sample Date/Time: Wednesday, June 29, 2016 12:31:24
Number of Replicates: 3
Autosampler Position: 307
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76244.7	1.6				ug/L	74699	Standard
	Be	9	10.0	50.0	**0.0187**	0.005	28.8	ug/L	8	Standard
	Al	27	1789441.8	2.2	**11.9153**	0.077	0.6	ug/L	1493	Standard
	Sc	45	61370.7	1.9				ug/L	61029	Standard
	Ti	47	59.7	11.8	**0.0448**	0.018	40.0	ug/L	35	Standard
	V	51	1976.3	1.1	**0.0601**	0.005	8.2	ug/L	1435	Standard
	Cr	52	8828.6	1.1	**0.1403**	0.016	11.5	ug/L	7635	Standard
	Cr	53	945.0	0.5	**0.6692**	0.006	1.0	ug/L	363	Standard
	Mn	55	180374.5	2.7	**25.1617**	0.439	1.7	ug/L	747	Standard
	Co	59	619.0	3.1	**0.0543**	0.004	7.0	ug/L	234	Standard
	Ni	60	405.7	6.0	**0.0878**	0.011	12.8	ug/L	215	Standard
	Cu	65	208.0	10.3	**0.0056**	0.012	222.0	ug/L	104	Standard
	Zn	66	6989.3	1.3	**5.7979**	0.130	2.2	ug/L	363	Standard
>	**Ge**	72	506777.4	1.1				ug/L	495142	Standard
	As	75	17.4	166.6	**0.0747**	0.025	33.6	ug/L	-50	Standard
	Se	82	43.0	5.9	**0.2447**	0.022	8.9	ug/L	14	Standard
	Se-1	77	89.0	9.0	**0.5677**	0.097	17.0	ug/L	42	Standard
>	**Ga**	71	55.0	18.2				mg/L	42	Standard
	Rb	85	226.7	20.8				ug/L	17	Standard
	Y	89	430230.4	1.6				ug/L	417098	Standard
>	**Rh**	103	23.3	12.4				ug/L	13	Standard
	Mo	98	43.4	8.2	**0.0107**	0.002	14.5	ug/L	11	Standard
	Ag	107	80.3	9.2	**-0.0005**	0.001	248.5	ug/L	68	Standard
	Cd	111	11.9	16.8	**-0.0022**	0.001	43.2	mg/L	2	Standard
	Cd	114	28.1	50.1	**0.0034**	0.003	82.8	ug/L	12	Standard
>	**In**	115	584282.7	1.3				ug/L	587441	Standard
	Sn	118	445.0	7.0	**-0.0181**	0.005	24.8	ug/L	497	Standard
	Sb	123	113.1	24.2	**0.0201**	0.006	28.5	ug/L	78	Standard
	Ba	135	83408.3	1.0	**40.9596**	0.601	1.5	ug/L	24	Standard
	Ce	140	301.7	22.6				ug/L	25	Standard
>	**Tb**	159	946714.8	0.6				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	182.7	3.3	**0.0205**	0.001	3.9	ug/L	4	Standard
	Tl	205	160.0	30.8	**0.0227**	0.007	30.9	ug/L	2	Standard
	Pb	206	288.0	1.3	**-0.0170**	0.001	5.3	ug/L	240	Standard
	Pb	207	270.7	6.0	**-0.0126**	0.003	25.4	ug/L	198	Standard
	Pb	208	1097.0	10.5	**-0.0153**	0.006	38.6	ug/L	883	Standard
	U	238	866.7	1.0	**0.1222**	0.001	0.8	ug/L	3	Standard
>	**Bi**	209	520921.7	0.5				ug/L	523984	Standard

Sample ID: L1606082303
Report Date/Time: Wednesday, June 29, 2016 12:33:40
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.3655**	33.533	173.2	mg/L	0	Standard
Mg	24	4915.8	1.0	**6.2466**	0.131	2.1	mg/L	37	Standard
K	39	50.0	20.0	**-0.0042**	0.023	541.6	mg/L	38	Standard
Ca	43	65.0	0.0	**1.5606**	0.079	5.0	mg/L	10	Standard
Fe	54	160.0	11.0	**-0.0035**	0.013	359.3	mg/L	179	Standard
Fe	57	281.7	9.1	**-0.1230**	0.068	55.3	mg/L	335	Standard
Sc-1	45	61370.7	1.9				mg/L	61029	Standard
Cl	35	23872.8	0.3				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	8675.8	8.1				ug/L	1547	Standard
P	31	16263.7	3.0				ug/L	15286	Standard
S	34	1945.1	6.6				ug/L	1572	Standard
Sr	88	130.0	24.0				ug/L	73	Standard
C	12	46.7	44.6				mg/L	63	Standard
N	14	6.7	86.6				mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	19.0	57.5				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	20.0	100.0				mg/L	13	Standard
I	127	19577.6	2.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		102.070	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		102.350	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606082303
Report Date/Time: Wednesday, June 29, 2016 12:33:40
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.462
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.415
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082304
Sample Date/Time: Wednesday, June 29, 2016 12:34:35
Number of Replicates: 3
Autosampler Position: 308
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76966.7	2.1				ug/L	74699	Standard
	Be	9	8.3	91.7	**0.0169**	0.009	50.5	ug/L	8	Standard
	Al	27	441491.6	0.4	**2.9030**	0.060	2.1	ug/L	1493	Standard
	Sc	45	60155.9	2.7				ug/L	61029	Standard
	Ti	47	44.0	12.0	**0.0058**	0.013	219.4	ug/L	35	Standard
	V	51	1675.2	9.7	**0.0239**	0.019	79.8	ug/L	1435	Standard
	Cr	52	8137.2	4.7	**0.0630**	0.039	62.0	ug/L	7635	Standard
	Cr	53	611.7	11.7	**0.3248**	0.069	21.3	ug/L	363	Standard
	Mn	55	105254.7	1.8	**14.8182**	0.156	1.1	ug/L	747	Standard
	Co	59	259.7	5.8	**0.0017**	0.003	156.1	ug/L	234	Standard
	Ni	60	230.0	4.9	**-0.0022**	0.007	321.8	ug/L	215	Standard
	Cu	65	165.3	1.5	**-0.0157**	0.002	12.9	ug/L	104	Standard
	Zn	66	2212.5	1.5	**1.5507**	0.010	0.6	ug/L	363	Standard
>	**Ge**	72	500132.0	1.1				ug/L	495142	Standard
	As	75	21.8	38.4	**0.0789**	0.008	9.6	ug/L	-50	Standard
	Se	82	26.9	8.6	**0.1047**	0.023	22.2	ug/L	14	Standard
	Se-1	77	64.0	26.1	**0.2460**	0.230	93.4	ug/L	42	Standard
>	**Ga**	71	55.0	41.7				mg/L	42	Standard
	Rb	85	421.7	18.5				ug/L	17	Standard
	Y	89	429897.2	1.0				ug/L	417098	Standard
>	**Rh**	103	6.7	43.3				ug/L	13	Standard
	Mo	98	29.2	13.0	**0.0052**	0.001	25.4	ug/L	11	Standard
	Ag	107	87.7	6.3	**0.0005**	0.001	182.2	ug/L	68	Standard
	Cd	111	7.0	28.7	**-0.0046**	0.001	19.4	mg/L	2	Standard
	Cd	114	21.6	74.7	**0.0021**	0.003	150.8	ug/L	12	Standard
>	**In**	115	589054.0	1.8				ug/L	587441	Standard
	Sn	118	438.3	5.7	**-0.0199**	0.003	16.6	ug/L	497	Standard
	Sb	123	106.2	8.9	**0.0185**	0.002	12.5	ug/L	78	Standard
	Ba	135	25689.8	1.2	**12.5057**	0.379	3.0	ug/L	24	Standard
	Ce	140	65.0	26.6				ug/L	25	Standard
>	**Tb**	159	953435.2	1.7				ug/L	939701	Standard
	Ho	165	13.3	43.3				ug/L	10	Standard
	Tl	203	91.7	10.9	**0.0091**	0.001	12.8	ug/L	4	Standard
	Tl	205	101.7	11.4	**0.0141**	0.002	11.0	ug/L	2	Standard
	Pb	206	281.0	2.2	**-0.0197**	0.001	5.9	ug/L	240	Standard
	Pb	207	219.0	9.1	**-0.0248**	0.004	16.3	ug/L	198	Standard
	Pb	208	999.0	3.0	**-0.0219**	0.001	6.3	ug/L	883	Standard
	U	238	2.7	21.7	**-0.0040**	0.000	2.0	ug/L	3	Standard
>	**Bi**	209	535313.4	0.3				ug/L	523984	Standard

Sample ID: L1606082304
Report Date/Time: Wednesday, June 29, 2016 12:36:52
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	1101.7	8.8	**1.3766**	0.099	7.2	mg/L	37	Standard
K	39	50.0	26.5	**-0.0017**	0.031	1783.7	mg/L	38	Standard
Ca	43	43.3	17.6	**0.2622**	0.555	211.4	mg/L	10	Standard
Fe	54	1079.1	4.5	**0.7903**	0.045	5.7	mg/L	179	Standard
Fe	57	550.0	4.8	**0.8245**	0.134	16.2	mg/L	335	Standard
> **Sc-1**	45	60155.9	2.7				mg/L	61029	Standard
Cl	35	23751.9	0.9				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	3913.8	0.4				ug/L	1547	Standard
P	31	16552.4	4.0				ug/L	15286	Standard
S	34	1948.5	5.9				ug/L	1572	Standard
Sr	88	68.3	30.5				ug/L	73	Standard
C	12	63.3	36.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	19.7	135.8				mg/L	6	Standard
Ho-1	165	13.3	43.3				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	5207.6	8.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.036	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.008	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082304
Report Date/Time: Wednesday, June 29, 2016 12:36:52
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.275
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.162
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 12:37:48
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	72653.8	1.6				ug/L	74699	Standard
	Be	9	42693.4	2.4	**49.8778**	1.581	3.2	ug/L	8	Standard
	Al	27	7018269.3	1.5	**49.0877**	0.211	0.4	ug/L	1493	Standard
	Sc	45	58700.2	1.7				ug/L	61029	Standard
	Ti	47	37360.7	1.5	**96.4941**	1.122	1.2	ug/L	35	Standard
	V	51	389361.0	0.6	**50.3345**	0.950	1.9	ug/L	1435	Standard
	Cr	52	377308.6	0.5	**48.8761**	0.835	1.7	ug/L	7635	Standard
	Cr	53	46063.3	3.3	**48.5325**	2.229	4.6	ug/L	363	Standard
	Mn	55	355991.7	2.4	**49.9570**	0.746	1.5	ug/L	747	Standard
	Co	59	341523.2	1.7	**50.6452**	0.528	1.0	ug/L	234	Standard
	Ni	60	94080.5	0.6	**49.1230**	0.490	1.0	ug/L	215	Standard
	Cu	65	95078.4	0.9	**49.9689**	1.178	2.4	ug/L	104	Standard
	Zn	66	56040.5	0.8	**49.8274**	0.722	1.4	ug/L	363	Standard
>	Ge	72	505267.2	1.6				ug/L	495142	Standard
	As	75	56871.7	1.3	**49.8175**	0.372	0.7	ug/L	-50	Standard
	Se	82	5593.2	2.1	**49.8004**	0.347	0.7	ug/L	14	Standard
	Se-1	77	3775.1	1.7	**50.0311**	0.568	1.1	ug/L	42	Standard
>	Ga	71	75.0	13.3				mg/L	42	Standard
	Rb	85	800.0	10.3				ug/L	17	Standard
	Y	89	408700.7	0.8				ug/L	417098	Standard
>	Rh	103	18.3	41.7				ug/L	13	Standard
	Mo	98	257910.1	1.9	**100.4657**	4.126	4.1	ug/L	11	Standard
	Ag	107	316989.6	1.3	**50.4692**	0.906	1.8	ug/L	68	Standard
	Cd	111	101914.0	1.0	**50.5948**	1.545	3.1	mg/L	2	Standard
	Cd	114	242699.9	2.4	**49.1311**	2.013	4.1	ug/L	12	Standard
>	In	115	566154.7	2.7				ug/L	587441	Standard
	Sn	118	281464.1	1.7	**49.9153**	1.113	2.2	ug/L	497	Standard
	Sb	123	227316.6	0.8	**50.5420**	1.534	3.0	ug/L	78	Standard
	Ba	135	98379.4	0.9	**49.8785**	1.233	2.5	ug/L	24	Standard
	Ce	140	113.3	20.8				ug/L	25	Standard
>	Tb	159	932106.1	1.1				ug/L	939701	Standard
	Ho	165	8.3	91.7				ug/L	10	Standard
	Tl	203	406761.4	0.1	**49.7003**	0.911	1.8	ug/L	4	Standard
	Tl	205	354916.5	2.0	**49.9771**	1.877	3.8	ug/L	2	Standard
	Pb	206	263562.6	0.6	**49.6234**	0.874	1.8	ug/L	240	Standard
	Pb	207	235278.4	0.8	**49.8612**	1.016	2.0	ug/L	198	Standard
	Pb	208	929662.3	0.8	**49.8673**	1.247	2.5	ug/L	883	Standard
	U	238	336021.4	1.5	**49.1730**	1.579	3.2	ug/L	3	Standard
>	Bi	209	520331.8	1.7				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:40:05
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.5945**	33.930	173.2	mg/L	0	Standard
Mg	24	3858.8	4.6	**5.1144**	0.246	4.8	mg/L	37	Standard
K	39	2055.1	3.6	**4.6098**	0.203	4.4	mg/L	38	Standard
Ca	43	130.0	23.4	**5.9979**	1.898	31.6	mg/L	10	Standard
Fe	54	5737.9	3.9	**4.9227**	0.278	5.7	mg/L	179	Standard
Fe	57	1583.4	3.8	**4.5245**	0.301	6.7	mg/L	335	Standard
Sc-1	45	58700.2	1.7				mg/L	61029	Standard
Cl	35	23099.6	1.9				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	1760.1	10.8				ug/L	1547	Standard
P	31	17002.9	0.9				ug/L	15286	Standard
S	34	2305.2	8.6				ug/L	1572	Standard
Sr	88	76.7	10.0				ug/L	73	Standard
C	12	60.0	33.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	9.0	191.6				mg/L	6	Standard
Ho-1	165	8.3	91.7				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	2030.1	31.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	99.756		
Al	27	98.175		
Sc	45			
Ti	47	96.494		
V	51	100.669		
Cr	52	97.752		
Cr	53			
Mn	55	99.914		
Co	59	101.290		
Ni	60	98.246		
Cu	65	99.938		
Zn	66	99.655		
Ge	72		102.045	
As	75	99.635		
Se	82	99.601		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:40:05
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	100.466		
Ag	107	100.938		
Cd	111	101.190		
Cd	114			
In	115			96.376
Sn	118	99.831		
Sb	123	101.084		
Ba	135	99.757		
Ce	140			
Tb	159			
Ho	165			
Tl	203	99.401		
Tl	205			
Pb	206			
Pb	207			
Pb	208	99.735		
U	238	98.346		
Bi	209			99.303
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:40:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 12:40:59
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73745.7	1.0				ug/L	74699	Standard
	Be	9	45.0	86.8	**0.0596**	0.045	76.2	ug/L	8	Standard
	Al	27	10763.7	138.2	**0.0612**	0.103	168.8	ug/L	1493	Standard
	Sc	45	59794.4	0.5				ug/L	61029	Standard
	Ti	47	63.7	40.2	**0.0565**	0.070	123.9	ug/L	35	Standard
	V	51	1309.0	5.1	**-0.0256**	0.009	36.2	ug/L	1435	Standard
	Cr	52	7144.0	4.7	**-0.0780**	0.064	81.5	ug/L	7635	Standard
	Cr	53	391.7	5.2	**0.0859**	0.030	34.5	ug/L	363	Standard
	Mn	55	2203.3	114.3	**0.1667**	0.364	218.6	ug/L	747	Standard
	Co	59	375.0	36.8	**0.0187**	0.022	116.6	ug/L	234	Standard
	Ni	60	226.7	25.3	**-0.0048**	0.033	679.9	ug/L	215	Standard
	Cu	65	137.7	45.7	**-0.0307**	0.035	113.4	ug/L	104	Standard
	Zn	66	391.7	8.9	**-0.1025**	0.038	37.5	ug/L	363	Standard
>	**Ge**	72	505094.9	1.9				ug/L	495142	Standard
	As	75	-30.0	135.0	**0.0336**	0.035	105.0	ug/L	-50	Standard
	Se	82	18.1	21.6	**0.0240**	0.035	145.0	ug/L	14	Standard
	Se-1	77	51.7	16.7	**0.0700**	0.104	148.9	ug/L	42	Standard
>	**Ga**	71	55.0	27.3				mg/L	42	Standard
	Rb	85	30.0	72.6				ug/L	17	Standard
	Y	89	416389.1	0.5				ug/L	417098	Standard
>	**Rh**	103	18.3	56.8				ug/L	13	Standard
	Mo	98	292.6	24.5	**0.1060**	0.029	27.7	ug/L	11	Standard
	Ag	107	123.7	36.8	**0.0063**	0.007	108.0	ug/L	68	Standard
	Cd	111	19.1	60.5	**0.0013**	0.005	411.5	mg/L	2	Standard
	Cd	114	51.4	75.2	**0.0080**	0.007	93.3	ug/L	12	Standard
>	**In**	115	578912.4	1.7				ug/L	587441	Standard
	Sn	118	756.7	23.6	**0.0364**	0.029	79.3	ug/L	497	Standard
	Sb	123	391.4	30.7	**0.0806**	0.025	30.9	ug/L	78	Standard
	Ba	135	64.3	74.2	**0.0157**	0.023	147.1	ug/L	24	Standard
	Ce	140	98.3	142.4				ug/L	25	Standard
>	**Tb**	159	938659.9	2.1				ug/L	939701	Standard
	Ho	165	16.7	45.8				ug/L	10	Standard
	Tl	203	80.3	33.4	**0.0079**	0.003	39.1	ug/L	4	Standard
	Tl	205	263.3	142.0	**0.0362**	0.051	140.6	ug/L	2	Standard
	Pb	206	290.0	20.9	**-0.0175**	0.011	60.1	ug/L	240	Standard
	Pb	207	255.0	4.1	**-0.0168**	0.002	14.3	ug/L	198	Standard
	Pb	208	1806.4	80.0	**0.0207**	0.074	360.4	ug/L	883	Standard
	U	238	234.3	149.2	**0.0289**	0.049	171.5	ug/L	3	Standard
>	**Bi**	209	529175.1	1.6				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:43:16
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	48.3	47.8	**-0.0014**	0.030	2108.5	mg/L	37	Standard
K	39	68.3	15.2	**0.0400**	0.023	56.9	mg/L	38	Standard
Ca	43	21.7	74.2	**-1.1227**	1.047	93.2	mg/L	10	Standard
Fe	54	169.2	23.1	**0.0081**	0.034	414.8	mg/L	179	Standard
Fe	57	310.0	16.4	**0.0019**	0.182	9789.6	mg/L	335	Standard
Sc-1	45	59794.4	0.5				mg/L	61029	Standard
Cl	35	23378.7	0.4				ug/L	20195	Standard
Kr	83	4.3	58.1				ug/L	4	Standard
Br	81	1630.1	3.1				ug/L	1547	Standard
P	31	16767.6	2.6				ug/L	15286	Standard
S	34	2153.5	8.8				ug/L	1572	Standard
Sr	88	65.0	27.7				ug/L	73	Standard
C	12	66.7	17.3				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	9.4	186.5				mg/L	6	Standard
Ho-1	165	16.7	45.8				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	1680.1	29.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		102.010	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:43:16
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	98.548	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	100.991	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:43:16
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Wednesday, June 29, 2016 12:44:12
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	71218.7	0.8				ug/L	74699	Standard
	Be	9	178.3	13.3	**0.2201**	0.029	13.2	ug/L	8	Standard
	Al	27	1766.8	29.5	**-0.0008**	0.004	458.8	ug/L	1493	Standard
	Sc	45	56025.1	1.1				ug/L	61029	Standard
	Ti	47	28.0	7.1	**-0.0318**	0.007	20.5	ug/L	35	Standard
	V	51	4079.9	5.1	**0.3696**	0.041	11.2	ug/L	1435	Standard
	Cr	52	12699.0	0.7	**0.7676**	0.041	5.4	ug/L	7635	Standard
	Cr	53	1050.0	6.6	**0.8579**	0.085	9.9	ug/L	363	Standard
	Mn	55	4110.2	0.9	**0.4692**	0.021	4.5	ug/L	747	Standard
	Co	59	2790.3	0.6	**0.4045**	0.014	3.5	ug/L	234	Standard
	Ni	60	2962.3	1.6	**1.5299**	0.023	1.5	ug/L	215	Standard
	Cu	65	1521.1	1.5	**0.7509**	0.034	4.5	ug/L	104	Standard
	Zn	66	7672.3	2.1	**6.8861**	0.035	0.5	ug/L	363	Standard
>	**Ge**	72	473802.8	2.6				ug/L	495142	Standard
	As	75	358.5	4.9	**0.3943**	0.021	5.3	ug/L	-50	Standard
	Se	82	48.3	16.1	**0.3230**	0.086	26.5	ug/L	14	Standard
	Se-1	77	78.0	9.0	**0.4927**	0.075	15.2	ug/L	42	Standard
>	**Ga**	71	35.0	24.7				mg/L	42	Standard
	Rb	85	21.7	35.3				ug/L	17	Standard
	Y	89	406616.3	2.6				ug/L	417098	Standard
>	**Rh**	103	13.3	43.3				ug/L	13	Standard
	Mo	98	80.8	10.7	**0.0260**	0.003	11.1	ug/L	11	Standard
	Ag	107	2461.5	2.7	**0.3827**	0.004	1.0	ug/L	68	Standard
	Cd	111	469.2	3.6	**0.2272**	0.006	2.5	mg/L	2	Standard
	Cd	114	1121.7	10.1	**0.2273**	0.026	11.5	ug/L	12	Standard
>	**In**	115	560717.1	1.7				ug/L	587441	Standard
	Sn	118	496.7	12.6	**-0.0055**	0.012	217.4	ug/L	497	Standard
	Sb	123	1804.7	2.2	**0.4010**	0.016	3.9	ug/L	78	Standard
	Ba	135	1433.1	2.6	**0.7176**	0.020	2.8	ug/L	24	Standard
	Ce	140	53.3	43.3				ug/L	25	Standard
>	**Tb**	159	901191.9	2.3				ug/L	939701	Standard
	Ho	165	5.0	0.0				ug/L	10	Standard
	Tl	203	658.3	6.7	**0.0809**	0.005	6.0	ug/L	4	Standard
	Tl	205	580.0	6.7	**0.0841**	0.005	5.7	ug/L	2	Standard
	Pb	206	1314.1	4.2	**0.1836**	0.008	4.1	ug/L	240	Standard
	Pb	207	1090.4	5.0	**0.1679**	0.009	5.1	ug/L	198	Standard
	Pb	208	4577.3	1.3	**0.1788**	0.005	3.0	ug/L	883	Standard
	U	238	2625.2	2.7	**0.3906**	0.007	1.9	ug/L	3	Standard
>	**Bi**	209	505831.0	1.4				ug/L	523984	Standard

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 12:46:29
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	40.0	12.5	**-0.0088**	0.007	78.9	mg/L	37	Standard
K	39	38.3	19.9	**-0.0218**	0.018	84.6	mg/L	38	Standard
Ca	43	36.7	75.1	**-0.0119**	1.877	15817.5	mg/L	10	Standard
Fe	54	97.1	24.3	**-0.0484**	0.023	47.3	mg/L	179	Standard
Fe	57	295.0	11.9	**0.0175**	0.123	699.1	mg/L	335	Standard
> **Sc-1**	45	56025.1	1.1				mg/L	61029	Standard
Cl	35	22767.8	1.3				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	1303.4	5.0				ug/L	1547	Standard
P	31	11956.4	1.3				ug/L	15286	Standard
S	34	2061.8	6.1				ug/L	1572	Standard
Sr	88	80.0	22.5				ug/L	73	Standard
C	12	80.0	33.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	0.0					mg/L	6	Standard
Ho-1	165	5.0	0.0				mg/L	10	Standard
Er	166	0.0					mg/L	13	Standard
I	127	1360.1	35.3				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	110.075		
	Al	27			
	Sc	45			
	Ti	47			
	V	51	92.400		
	Cr	52	95.955		
	Cr	53			
	Mn	55	93.844		
	Co	59	101.134		
	Ni	60	95.616		
	Cu	65	93.867		
	Zn	66	110.178		
>	Ge	72		95.690	
	As	75	98.586		
	Se	82	80.738		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 12:46:29
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	95.685		
Cd	111	94.646		
Cd	114			
> In	115			95.451
Sn	118			
Sb	123	100.239		
Ba	135	95.683		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.186		
Tl	205			
Pb	206			
Pb	207			
Pb	208	89.410		
U	238	97.655		
> Bi	209			96.535
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 12:46:29
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBW 7 WG573548-02
Sample Date/Time: Wednesday, June 29, 2016 12:53:54
Number of Replicates: 3
Autosampler Position: 309
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74857.8	1.6				ug/L	74699	Standard
	Be	9	16.7	75.5	**0.0264**	0.014	53.3	ug/L	8	Standard
	Al	27	80794.9	2.0	**0.5353**	0.014	2.5	ug/L	1493	Standard
	Sc	45	58461.0	3.0				ug/L	61029	Standard
	Ti	47	78.0	5.9	**0.0921**	0.013	14.5	ug/L	35	Standard
	V	51	1804.8	0.8	**0.0380**	0.004	11.0	ug/L	1435	Standard
	Cr	52	9792.5	0.9	**0.2676**	0.021	7.7	ug/L	7635	Standard
	Cr	53	633.3	5.7	**0.3397**	0.038	11.1	ug/L	363	Standard
	Mn	55	7868.0	0.9	**0.9563**	0.019	2.0	ug/L	747	Standard
	Co	59	389.7	6.7	**0.0204**	0.003	16.2	ug/L	234	Standard
	Ni	60	396.3	4.2	**0.0829**	0.008	9.4	ug/L	215	Standard
	Cu	65	182.0	8.2	**-0.0081**	0.009	109.2	ug/L	104	Standard
	Zn	66	1432.4	5.9	**0.8274**	0.087	10.5	ug/L	363	Standard
>	**Ge**	72	506709.7	1.0				ug/L	495142	Standard
	As	75	61.7	54.0	**0.1136**	0.029	25.8	ug/L	-50	Standard
	Se	82	57.7	7.0	**0.3762**	0.040	10.6	ug/L	14	Standard
	Se-1	77	48.3	5.2	**0.0243**	0.031	128.3	ug/L	42	Standard
>	**Ga**	71	106.7	35.5				mg/L	42	Standard
	Rb	85	413.3	4.9				ug/L	17	Standard
	Y	89	414741.4	0.4				ug/L	417098	Standard
>	**Rh**	103	15.0	100.0				ug/L	13	Standard
	Mo	98	46.5	23.9	**0.0125**	0.004	33.5	ug/L	11	Standard
	Ag	107	1549.7	3.3	**0.2354**	0.012	5.1	ug/L	68	Standard
	Cd	111	17.2	8.9	**0.0007**	0.001	113.7	mg/L	2	Standard
	Cd	114	63.6	36.2	**0.0109**	0.005	44.4	ug/L	12	Standard
>	**In**	115	562869.1	1.7				ug/L	587441	Standard
	Sn	118	736.7	11.6	**0.0370**	0.017	44.6	ug/L	497	Standard
	Sb	123	216.9	10.9	**0.0443**	0.006	12.6	ug/L	78	Standard
	Ba	135	906.0	5.8	**0.4461**	0.030	6.6	ug/L	24	Standard
	Ce	140	1735.1	5.3				ug/L	25	Standard
>	**Tb**	159	918530.3	1.8				ug/L	939701	Standard
	Ho	165	31.7	32.9				ug/L	10	Standard
	Tl	203	89.3	15.5	**0.0092**	0.002	18.3	ug/L	4	Standard
	Tl	205	86.7	13.3	**0.0124**	0.002	13.0	ug/L	2	Standard
	Pb	206	378.0	4.1	**0.0001**	0.003	2503.4	ug/L	240	Standard
	Pb	207	310.7	3.7	**-0.0040**	0.002	56.8	ug/L	198	Standard
	Pb	208	1268.7	1.9	**-0.0059**	0.001	22.1	ug/L	883	Standard
	U	238	52.0	18.9	**0.0032**	0.001	44.2	ug/L	3	Standard
>	**Bi**	209	520017.2	0.7				ug/L	523984	Standard

Sample ID: PBW 7 WG573548-02
Report Date/Time: Wednesday, June 29, 2016 12:56:11
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	98.3	32.7	**0.0678**	0.045	67.0	mg/L	37	Standard
K	39	61.7	9.4	**0.0281**	0.012	42.4	mg/L	38	Standard
Ca	43	31.7	55.5	**-0.4154**	1.202	289.4	mg/L	10	Standard
Fe	54	153.7	8.4	**-0.0024**	0.007	308.5	mg/L	179	Standard
Fe	57	316.7	15.8	**0.0490**	0.173	353.8	mg/L	335	Standard
> **Sc-1**	45	58461.0	3.0				mg/L	61029	Standard
Cl	35	22819.8	1.0				ug/L	20195	Standard
Kr	83	3.7	31.5				ug/L	4	Standard
Br	81	12281.6	4.1				ug/L	1547	Standard
P	31	15094.2	2.5				ug/L	15286	Standard
S	34	1958.5	7.0				ug/L	1572	Standard
Sr	88	78.3	19.5				ug/L	73	Standard
C	12	96.7	21.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	48.1	42.1				mg/L	6	Standard
Ho-1	165	31.7	32.9				mg/L	10	Standard
Er	166	40.0	25.0				mg/L	13	Standard
I	127	2060.1	27.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		100.213	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.336	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBW 7 WG573548-02
Report Date/Time: Wednesday, June 29, 2016 12:56:11
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	95.817	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	99.243	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 7 WG573548-02
Report Date/Time: Wednesday, June 29, 2016 12:56:11
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSW 7 WG573548-03
Sample Date/Time: Wednesday, June 29, 2016 12:57:05
Number of Replicates: 3
Autosampler Position: 310
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75779.1	1.4				ug/L	74699	Standard
	Be	9	44266.3	2.0	**49.5806**	1.413	2.8	ug/L	8	Standard
	Al	27	95277.1	2.2	**0.6259**	0.022	3.5	ug/L	1493	Standard
	Sc	45	60008.6	1.7				ug/L	61029	Standard
	Ti	47	61.3	19.5	**0.0533**	0.033	61.1	ug/L	35	Standard
	V	51	383994.5	1.5	**50.7641**	0.334	0.7	ug/L	1435	Standard
	Cr	52	391642.5	1.8	**51.9470**	0.996	1.9	ug/L	7635	Standard
	Cr	53	48176.5	1.3	**51.9174**	0.408	0.8	ug/L	363	Standard
	Mn	55	379234.4	0.1	**54.4505**	0.406	0.7	ug/L	747	Standard
	Co	59	337441.4	0.9	**51.1841**	0.650	1.3	ug/L	234	Standard
	Ni	60	96944.2	1.6	**51.7745**	0.652	1.3	ug/L	215	Standard
	Cu	65	97447.4	0.9	**52.3759**	0.338	0.6	ug/L	104	Standard
	Zn	66	58710.9	1.4	**53.4195**	0.683	1.3	ug/L	363	Standard
>	**Ge**	72	493989.7	0.8				ug/L	495142	Standard
	As	75	54420.0	2.0	**48.7560**	0.745	1.5	ug/L	-50	Standard
	Se	82	5499.8	2.1	**50.0913**	1.110	2.2	ug/L	14	Standard
	Se-1	77	3757.5	2.3	**50.9460**	1.277	2.5	ug/L	42	Standard
>	**Ga**	71	331.7	6.3				mg/L	42	Standard
	Rb	85	1905.1	7.7				ug/L	17	Standard
	Y	89	430193.5	2.2				ug/L	417098	Standard
>	**Rh**	103	21.7	13.3				ug/L	13	Standard
	Mo	98	200.3	5.0	**0.0686**	0.004	5.6	ug/L	11	Standard
	Ag	107	325181.2	0.5	**49.3255**	0.456	0.9	ug/L	68	Standard
	Cd	111	102382.7	0.7	**48.4138**	0.589	1.2	mg/L	2	Standard
	Cd	114	246010.3	1.2	**47.4297**	0.453	1.0	ug/L	12	Standard
>	**In**	115	594099.2	0.9				ug/L	587441	Standard
	Sn	118	733.4	7.5	**0.0295**	0.010	35.4	ug/L	497	Standard
	Sb	123	222790.7	0.6	**47.1842**	0.681	1.4	ug/L	78	Standard
	Ba	135	100456.3	0.7	**48.5175**	0.574	1.2	ug/L	24	Standard
	Ce	140	3297.0	2.9				ug/L	25	Standard
>	**Tb**	159	953975.0	2.3				ug/L	939701	Standard
	Ho	165	81.7	3.5				ug/L	10	Standard
	Tl	203	411383.7	1.0	**49.0691**	0.983	2.0	ug/L	4	Standard
	Tl	205	358822.0	2.1	**49.3118**	1.192	2.4	ug/L	2	Standard
	Pb	206	272876.4	0.6	**50.1549**	0.805	1.6	ug/L	240	Standard
	Pb	207	232281.8	1.2	**48.0517**	0.976	2.0	ug/L	198	Standard
	Pb	208	938885.2	1.2	**49.1601**	1.120	2.3	ug/L	883	Standard
	U	238	337957.9	0.3	**48.2703**	0.691	1.4	ug/L	3	Standard
>	**Bi**	209	532979.3	1.2				ug/L	523984	Standard

Sample ID: LCSW 7 WG573548-03
Report Date/Time: Wednesday, June 29, 2016 12:59:22
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	40.0	12.5	**-0.0126**	0.007	52.7	mg/L	37	Standard
K	39	58.3	21.6	**0.0173**	0.030	174.9	mg/L	38	Standard
Ca	43	60.0	16.7	**1.3313**	0.636	47.7	mg/L	10	Standard
Fe	54	427.3	5.7	**0.2300**	0.018	7.6	mg/L	179	Standard
Fe	57	393.3	4.8	**0.2864**	0.089	31.1	mg/L	335	Standard
> **Sc-1**	45	60008.6	1.7				mg/L	61029	Standard
Cl	35	22904.0	1.4				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	1933.5	11.4				ug/L	1547	Standard
P	31	15693.1	2.0				ug/L	15286	Standard
S	34	1913.5	5.4				ug/L	1572	Standard
Sr	88	85.0	21.2				ug/L	73	Standard
C	12	106.7	37.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	149.3	43.4				mg/L	6	Standard
Ho-1	165	81.7	3.5				mg/L	10	Standard
Er	166	83.3	45.4				mg/L	13	Standard
I	127	1690.1	32.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.446	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.767	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 7 WG573548-03
Report Date/Time: Wednesday, June 29, 2016 12:59:22
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
In	115		101.133
Sn	118		
Sb	123		
Ba	135		
Ce	140		
Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
Bi	209		101.717
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 7 WG573548-03
Report Date/Time: Wednesday, June 29, 2016 12:59:22
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098623 WG573548-01
Sample Date/Time: Wednesday, June 29, 2016 13:00:17
Number of Replicates: 3
Autosampler Position: 311
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85536.1	4.6				ug/L	74699	Standard
	Be	9	131.7	4.4	**0.1385**	0.009	6.7	ug/L	8	Standard
	Al	27	51229429.5	1.9	**304.6734**	8.454	2.8	ug/L	1493	Standard
	Sc	45	58156.4	0.7				ug/L	61029	Standard
	Ti	47	291.7	4.0	**0.6976**	0.027	3.9	ug/L	35	Standard
	V	51	2731.2	8.5	**0.1837**	0.035	19.3	ug/L	1435	Standard
	Cr	52	11798.6	2.2	**0.6453**	0.034	5.3	ug/L	7635	Standard
	Cr	53	1541.7	10.7	**1.4199**	0.200	14.1	ug/L	363	Standard
	Mn	55	1519614.4	1.3	**228.6645**	1.850	0.8	ug/L	747	Standard
	Co	59	1958.8	0.9	**0.2737**	0.000	0.1	ug/L	234	Standard
	Ni	60	5664.7	1.6	**3.0483**	0.071	2.3	ug/L	215	Standard
	Cu	65	19925.4	0.1	**11.1201**	0.082	0.7	ug/L	104	Standard
	Zn	66	6263.0	2.6	**5.5559**	0.126	2.3	ug/L	363	Standard
>	**Ge**	72	472304.7	0.8				ug/L	495142	Standard
	As	75	721.9	3.2	**0.7353**	0.022	3.1	ug/L	-50	Standard
	Se	82	400.4	3.0	**3.6870**	0.142	3.9	ug/L	14	Standard
	Se-1	77	187.0	7.2	**2.0610**	0.173	8.4	ug/L	42	Standard
>	**Ga**	71	100.0	8.7				mg/L	42	Standard
	Rb	85	5192.6	2.9				ug/L	17	Standard
	Y	89	424150.6	2.5				ug/L	417098	Standard
>	**Rh**	103	291.7	21.8				ug/L	13	Standard
	Mo	98	1037.1	3.0	**0.3904**	0.016	4.1	ug/L	11	Standard
	Ag	107	150.3	11.8	**0.0106**	0.003	25.5	ug/L	68	Standard
	Cd	111	86.5	9.0	**0.0342**	0.004	11.5	mg/L	2	Standard
	Cd	114	235.1	23.0	**0.0445**	0.010	23.2	ug/L	12	Standard
>	**In**	115	577173.5	1.1				ug/L	587441	Standard
	Sn	118	1115.0	2.8	**0.0996**	0.004	4.3	ug/L	497	Standard
	Sb	123	1725.9	23.2	**0.3715**	0.084	22.5	ug/L	78	Standard
	Ba	135	22344.1	0.8	**11.0965**	0.201	1.8	ug/L	24	Standard
	Ce	140	1543.4	4.9				ug/L	25	Standard
>	**Tb**	159	936227.5	1.2				ug/L	939701	Standard
	Ho	165	78.3	35.2				ug/L	10	Standard
	Tl	203	240.0	4.8	**0.0287**	0.001	5.1	ug/L	4	Standard
	Tl	205	171.7	9.4	**0.0253**	0.002	9.3	ug/L	2	Standard
	Pb	206	666.7	6.9	**0.0592**	0.009	15.2	ug/L	240	Standard
	Pb	207	558.7	0.1	**0.0529**	0.000	0.3	ug/L	198	Standard
	Pb	208	2267.7	0.3	**0.0522**	0.000	0.9	ug/L	883	Standard
	U	238	28225.3	0.9	**4.2785**	0.042	1.0	ug/L	3	Standard
>	**Bi**	209	501665.1	0.1				ug/L	523984	Standard

Sample ID: L1606098623 WG573548-01
Report Date/Time: Wednesday, June 29, 2016 13:02:34
Page 1

Approved: July 04, 2016

Na	23	13.3	43.3	**161.5718**	71.131	44.0	mg/L	0	Standard
Mg	24	24900.4	1.1	**33.6647**	0.149	0.4	mg/L	37	Standard
K	39	1233.4	1.4	**2.7469**	0.062	2.2	mg/L	38	Standard
Ca	43	1560.1	5.5	**100.8648**	4.911	4.9	mg/L	10	Standard
Fe	54	167.1	25.8	**0.0105**	0.039	372.7	mg/L	179	Standard
Fe	57	958.4	6.3	**2.3440**	0.212	9.0	mg/L	335	Standard
Sc-1	45	58156.4	0.7				mg/L	61029	Standard
Cl	35	40465.3	2.5				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	58241.9	4.1				ug/L	1547	Standard
P	31	17453.4	1.6				ug/L	15286	Standard
S	34	1810.1	8.9				ug/L	1572	Standard
Sr	88	1313.4	8.1				ug/L	73	Standard
C	12	273.3	15.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	3	Standard
Dy	164	89.3	16.3				mg/L	6	Standard
Ho-1	165	78.3	35.2				mg/L	10	Standard
Er	166	83.3	18.3				mg/L	13	Standard
I	127	135078.2	7.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		114.508	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.388	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098623 WG573548-01
Report Date/Time: Wednesday, June 29, 2016 13:02:34
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.252
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.740
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098623 WG573548-01
Report Date/Time: Wednesday, June 29, 2016 13:02:34
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098635S WG573548-04
Sample Date/Time: Wednesday, June 29, 2016 13:03:28
Number of Replicates: 3
Autosampler Position: 312
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	96645.6	0.2				ug/L	74699	Standard
	Be	9	56332.9	2.6	**49.4619**	1.222	2.5	ug/L	8	Standard
	Al	27	56565131.5	1.0	**297.4859**	3.609	1.2	ug/L	1493	Standard
	Sc	45	58999.8	4.0				ug/L	61029	Standard
	Ti	47	369.0	11.5	**0.8968**	0.108	12.0	ug/L	35	Standard
	V	51	390242.6	0.5	**53.2266**	0.984	1.8	ug/L	1435	Standard
	Cr	52	371857.5	2.1	**50.8419**	0.408	0.8	ug/L	7635	Standard
	Cr	53	46974.4	2.3	**52.2114**	1.040	2.0	ug/L	363	Standard
	Mn	55	1851169.7	0.3	**274.7367**	3.654	1.3	ug/L	747	Standard
	Co	59	350457.1	0.5	**54.8326**	0.988	1.8	ug/L	234	Standard
	Ni	60	94372.4	2.0	**51.9777**	0.274	0.5	ug/L	215	Standard
	Cu	65	115580.4	2.6	**64.0828**	0.734	1.1	ug/L	104	Standard
	Zn	66	61118.4	0.5	**57.3902**	0.612	1.1	ug/L	363	Standard
>	Ge	72	478993.7	1.6				ug/L	495142	Standard
	As	75	59122.1	0.8	**54.6260**	0.454	0.8	ug/L	-50	Standard
	Se	82	6294.5	0.9	**59.1515**	0.578	1.0	ug/L	14	Standard
	Se-1	77	4101.2	2.7	**57.4162**	0.662	1.2	ug/L	42	Standard
>	Ga	71	100.0	18.0				mg/L	42	Standard
	Rb	85	5364.3	1.3				ug/L	17	Standard
	Y	89	423648.2	2.3				ug/L	417098	Standard
>	Rh	103	290.0	15.8				ug/L	13	Standard
	Mo	98	1231.1	6.2	**0.4537**	0.027	5.8	ug/L	11	Standard
	Ag	107	140053.7	1.3	**21.3673**	0.341	1.6	ug/L	68	Standard
	Cd	111	101619.5	1.1	**48.3438**	0.154	0.3	mg/L	2	Standard
	Cd	114	248812.6	0.3	**48.2667**	0.577	1.2	ug/L	12	Standard
>	In	115	590497.7	1.3				ug/L	587441	Standard
	Sn	118	1276.7	2.9	**0.1227**	0.007	5.6	ug/L	497	Standard
	Sb	123	230648.9	1.4	**49.1452**	0.703	1.4	ug/L	78	Standard
	Ba	135	121100.4	1.0	**58.8551**	1.255	2.1	ug/L	24	Standard
	Ce	140	1580.1	12.1				ug/L	25	Standard
>	Tb	159	967679.2	1.8				ug/L	939701	Standard
	Ho	165	70.0	12.4				ug/L	10	Standard
	Tl	203	411213.8	0.9	**50.0486**	0.144	0.3	ug/L	4	Standard
	Tl	205	371375.3	1.6	**52.0881**	1.342	2.6	ug/L	2	Standard
	Pb	206	271184.0	0.3	**50.8655**	0.565	1.1	ug/L	240	Standard
	Pb	207	230172.1	0.3	**48.5918**	0.740	1.5	ug/L	198	Standard
	Pb	208	934549.0	0.8	**49.9358**	0.874	1.8	ug/L	883	Standard
	U	238	386476.6	1.1	**56.3336**	1.024	1.8	ug/L	3	Standard
>	Bi	209	522269.6	1.2				ug/L	523984	Standard

Sample ID: L1606098635S WG573548-04
Report Date/Time: Wednesday, June 29, 2016 13:05:45
Page 1

Approved: July 04, 2016

Na	23	8.3	91.7	**99.7813**	89.869	90.1	mg/L	0	Standard
Mg	24	26446.4	1.6	**35.2703**	0.863	2.4	mg/L	37	Standard
K	39	1201.7	4.6	**2.6333**	0.063	2.4	mg/L	38	Standard
Ca	43	1648.4	4.4	**105.1755**	1.490	1.4	mg/L	10	Standard
Fe	54	207.2	35.2	**0.0431**	0.061	140.9	mg/L	179	Standard
Fe	57	976.7	6.7	**2.3580**	0.131	5.6	mg/L	335	Standard
Sc-1	45	58999.8	4.0				mg/L	61029	Standard
Cl	35	42664.0	0.9				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	62512.1	2.5				ug/L	1547	Standard
P	31	18094.1	2.0				ug/L	15286	Standard
S	34	1855.1	3.1				ug/L	1572	Standard
Sr	88	1376.7	4.2				ug/L	73	Standard
C	12	206.7	24.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	73.5	28.0				mg/L	6	Standard
Ho-1	165	70.0	12.4				mg/L	10	Standard
Er	166	66.7	34.6				mg/L	13	Standard
I	127	146421.6	4.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		129.380	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.739	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098635S WG573548-04
Report Date/Time: Wednesday, June 29, 2016 13:05:45
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	100.520
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	99.673
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098635S WG573548-04
Report Date/Time: Wednesday, June 29, 2016 13:05:45
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098637SD WG573548-05
Sample Date/Time: Wednesday, June 29, 2016 13:06:39
Number of Replicates: 3
Autosampler Position: 313
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	99478.9	1.8				ug/L	74699	Standard
	Be	9	55809.3	1.7	**47.6121**	0.679	1.4	ug/L	8	Standard
	Al	27	57079276.5	2.7	**291.6498**	6.801	2.3	ug/L	1493	Standard
	Sc	45	56220.8	1.7				ug/L	61029	Standard
	Ti	47	357.7	29.4	**0.8728**	0.310	35.5	ug/L	35	Standard
	V	51	392795.1	0.3	**53.7087**	1.385	2.6	ug/L	1435	Standard
	Cr	52	364211.2	1.2	**49.9063**	1.054	2.1	ug/L	7635	Standard
	Cr	53	45092.0	2.6	**50.2220**	0.967	1.9	ug/L	363	Standard
	Mn	55	1864825.2	1.1	**277.4410**	6.534	2.4	ug/L	747	Standard
	Co	59	357216.6	1.3	**56.0166**	0.665	1.2	ug/L	234	Standard
	Ni	60	94064.4	0.7	**51.9434**	1.099	2.1	ug/L	215	Standard
	Cu	65	96872.1	1.3	**53.8290**	0.581	1.1	ug/L	104	Standard
	Zn	66	56756.1	1.1	**53.3877**	0.715	1.3	ug/L	363	Standard
>	**Ge**	72	477920.9	2.4				ug/L	495142	Standard
	As	75	58834.8	0.8	**54.4924**	0.918	1.7	ug/L	-50	Standard
	Se	82	6411.7	0.6	**60.4029**	1.057	1.7	ug/L	14	Standard
	Se-1	77	4156.6	3.0	**58.3446**	1.540	2.6	ug/L	42	Standard
>	**Ga**	71	126.7	16.4				mg/L	42	Standard
	Rb	85	5344.3	1.7				ug/L	17	Standard
	Y	89	425761.1	0.9				ug/L	417098	Standard
>	**Rh**	103	338.3	14.4				ug/L	13	Standard
	Mo	98	943.7	3.1	**0.3564**	0.013	3.7	ug/L	11	Standard
	Ag	107	126054.2	0.9	**19.7696**	0.296	1.5	ug/L	68	Standard
	Cd	111	100507.1	1.6	**49.1528**	0.361	0.7	mg/L	2	Standard
	Cd	114	240725.5	3.1	**47.9933**	0.862	1.8	ug/L	12	Standard
>	**In**	115	574417.4	1.5				ug/L	587441	Standard
	Sn	118	1058.4	4.2	**0.0905**	0.005	5.8	ug/L	497	Standard
	Sb	123	229849.5	0.5	**50.3475**	0.456	0.9	ug/L	78	Standard
	Ba	135	119199.1	0.7	**59.5474**	0.534	0.9	ug/L	24	Standard
	Ce	140	1411.7	10.5				ug/L	25	Standard
>	**Tb**	159	946510.5	1.0				ug/L	939701	Standard
	Ho	165	61.7	32.8				ug/L	10	Standard
	Tl	203	404314.5	1.0	**50.2968**	0.720	1.4	ug/L	4	Standard
	Tl	205	368989.1	0.6	**52.8952**	1.157	2.2	ug/L	2	Standard
	Pb	206	274418.5	1.5	**52.6043**	0.331	0.6	ug/L	240	Standard
	Pb	207	233408.5	1.0	**50.3596**	0.394	0.8	ug/L	198	Standard
	Pb	208	945417.2	0.6	**51.6312**	0.704	1.4	ug/L	883	Standard
	U	238	390395.6	1.2	**58.1569**	0.854	1.5	ug/L	3	Standard
>	**Bi**	209	511024.4	1.7				ug/L	523984	Standard

Sample ID: L1606098637SD WG573548-05
Report Date/Time: Wednesday, June 29, 2016 13:08:56
Page 1

Approved: July 04, 2016

Na	23	13.3	21.7	**167.0389**	37.375	22.4	mg/L	0	Standard
Mg	24	26820.4	2.0	**37.5161**	0.222	0.6	mg/L	37	Standard
K	39	1165.0	5.8	**2.6799**	0.118	4.4	mg/L	38	Standard
Ca	43	1491.7	10.9	**99.7885**	11.256	11.3	mg/L	10	Standard
Fe	54	157.1	11.5	**0.0062**	0.015	238.1	mg/L	179	Standard
Fe	57	1011.7	5.7	**2.6573**	0.164	6.2	mg/L	335	Standard
Sc-1	45	56220.8	1.7				mg/L	61029	Standard
Cl	35	42892.6	1.5				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	62984.0	1.3				ug/L	1547	Standard
P	31	17081.3	3.0				ug/L	15286	Standard
S	34	1805.1	2.7				ug/L	1572	Standard
Sr	88	1383.4	10.5				ug/L	73	Standard
C	12	180.0	14.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	124.6	33.0				mg/L	6	Standard
Ho-1	165	61.7	32.8				mg/L	10	Standard
Er	166	43.3	35.3				mg/L	13	Standard
I	127	140810.4	4.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		133.173	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.522	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098637SD WG573548-05
Report Date/Time: Wednesday, June 29, 2016 13:08:56
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.783
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.527
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098637SD WG573548-05
Report Date/Time: Wednesday, June 29, 2016 13:08:56
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098601
Sample Date/Time: Wednesday, June 29, 2016 13:09:51
Number of Replicates: 3
Autosampler Position: 314
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	102324.4	1.8				ug/L	74699	Standard
	Be	9	88.3	32.2	**0.0810**	0.024	30.0	ug/L	8	Standard
	Al	27	5114148.7	1.9	**25.3952**	0.605	2.4	ug/L	1493	Standard
	Sc	45	60191.0	1.9				ug/L	61029	Standard
	Ti	47	2779.9	0.6	**7.0936**	0.108	1.5	ug/L	35	Standard
	V	51	5111.4	5.6	**0.4688**	0.033	7.0	ug/L	1435	Standard
	Cr	52	13420.3	3.5	**0.7542**	0.051	6.8	ug/L	7635	Standard
	Cr	53	1458.4	3.3	**1.2196**	0.051	4.2	ug/L	363	Standard
	Mn	55	181173.8	0.3	**25.4046**	0.180	0.7	ug/L	747	Standard
	Co	59	2348.5	2.0	**0.3119**	0.007	2.1	ug/L	234	Standard
	Ni	60	2310.5	3.1	**1.0878**	0.031	2.9	ug/L	215	Standard
	Cu	65	2308.5	3.6	**1.1141**	0.034	3.0	ug/L	104	Standard
	Zn	66	22662.3	0.5	**19.9177**	0.090	0.4	ug/L	363	Standard
>	**Ge**	72	504254.4	0.9				ug/L	495142	Standard
	As	75	859.1	5.1	**0.8126**	0.036	4.5	ug/L	-50	Standard
	Se	82	56.8	14.9	**0.3708**	0.079	21.2	ug/L	14	Standard
	Se-1	77	87.3	17.2	**0.5507**	0.190	34.5	ug/L	42	Standard
>	**Ga**	71	526.7	8.3				mg/L	42	Standard
	Rb	85	16517.3	5.3				ug/L	17	Standard
	Y	89	430302.4	1.6				ug/L	417098	Standard
>	**Rh**	103	86.7	33.8				ug/L	13	Standard
	Mo	98	1898.6	1.4	**0.7136**	0.022	3.1	ug/L	11	Standard
	Ag	107	115.3	8.0	**0.0050**	0.001	29.0	ug/L	68	Standard
	Cd	111	60.8	33.2	**0.0215**	0.010	47.6	mg/L	2	Standard
	Cd	114	128.6	20.8	**0.0232**	0.005	20.9	ug/L	12	Standard
>	**In**	115	581907.5	1.7				ug/L	587441	Standard
	Sn	118	1231.7	10.8	**0.1183**	0.024	20.7	ug/L	497	Standard
	Sb	123	520.8	14.7	**0.1084**	0.016	15.1	ug/L	78	Standard
	Ba	135	41346.0	0.9	**20.3791**	0.217	1.1	ug/L	24	Standard
	Ce	140	23020.8	2.4				ug/L	25	Standard
>	**Tb**	159	964144.3	1.8				ug/L	939701	Standard
	Ho	165	513.3	9.3				ug/L	10	Standard
	Tl	203	211.3	12.7	**0.0234**	0.003	14.7	ug/L	4	Standard
	Tl	205	253.3	62.3	**0.0350**	0.022	63.0	ug/L	2	Standard
	Pb	206	1500.1	2.6	**0.2039**	0.008	4.0	ug/L	240	Standard
	Pb	207	1190.0	4.0	**0.1756**	0.010	5.7	ug/L	198	Standard
	Pb	208	4936.0	2.0	**0.1838**	0.008	4.3	ug/L	883	Standard
	U	238	2034.5	10.8	**0.2852**	0.034	11.8	ug/L	3	Standard
>	**Bi**	209	535012.0	1.1				ug/L	523984	Standard

Sample ID: L1606098601
Report Date/Time: Wednesday, June 29, 2016 13:12:08
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	50263.2	0.4	**65.7389**	1.477	2.2	mg/L	37	Standard
K	39	250.0	7.2	**0.4466**	0.047	10.5	mg/L	38	Standard
Ca	43	233.3	8.9	**12.4104**	1.127	9.1	mg/L	10	Standard
Fe	54	825.5	3.8	**0.5718**	0.038	6.7	mg/L	179	Standard
Fe	57	516.7	4.4	**0.7058**	0.045	6.4	mg/L	335	Standard
> **Sc-1**	45	60191.0	1.9				mg/L	61029	Standard
Cl	35	28383.9	2.7				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	8439.0	2.4				ug/L	1547	Standard
P	31	17857.2	2.8				ug/L	15286	Standard
S	34	1791.8	4.6				ug/L	1572	Standard
Sr	88	506.7	4.7				ug/L	73	Standard
C	12	153.3	22.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	893.4	9.1				mg/L	6	Standard
Ho-1	165	513.3	9.3				mg/L	10	Standard
Er	166	416.7	13.2				mg/L	13	Standard
I	127	18778.3	7.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		136.983	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.840	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098601
Report Date/Time: Wednesday, June 29, 2016 13:12:08
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	99.058	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	102.105	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603
Sample Date/Time: Wednesday, June 29, 2016 13:13:02
Number of Replicates: 3
Autosampler Position: 315
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	106961.3	1.6				ug/L	74699	Standard
	Be	9	178.3	7.1	**0.1492**	0.012	8.2	ug/L	8	Standard
	Al	27	64490531.4	4.4	**306.3886**	9.864	3.2	ug/L	1493	Standard
	Sc	45	59811.2	2.1				ug/L	61029	Standard
	Ti	47	416.0	0.2	**0.9910**	0.025	2.5	ug/L	35	Standard
	V	51	3004.3	2.1	**0.2029**	0.003	1.2	ug/L	1435	Standard
	Cr	52	11901.7	1.5	**0.5854**	0.019	3.3	ug/L	7635	Standard
	Cr	53	991.7	2.3	**0.7459**	0.050	6.7	ug/L	363	Standard
	Mn	55	12370472.2	4.2	**1779.3602**	31.656	1.8	ug/L	747	Standard
	Co	59	18875.4	5.7	**2.8249**	0.100	3.5	ug/L	234	Standard
	Ni	60	12202.3	4.0	**6.4037**	0.114	1.8	ug/L	215	Standard
	Cu	65	16836.0	2.8	**8.9582**	0.039	0.4	ug/L	104	Standard
	Zn	66	24869.4	2.3	**22.3549**	0.039	0.2	ug/L	363	Standard
>	**Ge**	72	494237.2	2.4				ug/L	495142	Standard
	As	75	1554.4	6.7	**1.4489**	0.061	4.2	ug/L	-50	Standard
	Se	82	71.5	7.7	**0.5142**	0.037	7.1	ug/L	14	Standard
	Se-1	77	63.0	12.4	**0.2431**	0.119	49.1	ug/L	42	Standard
>	**Ga**	71	356.7	10.3				mg/L	42	Standard
	Rb	85	9309.5	2.8				ug/L	17	Standard
	Y	89	460268.8	4.1				ug/L	417098	Standard
>	**Rh**	103	310.0	9.7				ug/L	13	Standard
	Mo	98	1407.8	3.5	**0.5012**	0.005	1.0	ug/L	11	Standard
	Ag	107	101.3	22.9	**0.0020**	0.003	151.7	ug/L	68	Standard
	Cd	111	246.3	7.9	**0.1051**	0.008	7.2	mg/L	2	Standard
	Cd	114	605.6	11.0	**0.1111**	0.009	8.4	ug/L	12	Standard
>	**In**	115	612194.6	4.4				ug/L	587441	Standard
	Sn	118	1470.1	15.0	**0.1461**	0.027	18.7	ug/L	497	Standard
	Sb	123	727.1	5.4	**0.1452**	0.003	2.2	ug/L	78	Standard
	Ba	135	24533.2	4.5	**11.4855**	0.055	0.5	ug/L	24	Standard
	Ce	140	44839.6	2.7				ug/L	25	Standard
>	**Tb**	159	1002318.5	3.7				ug/L	939701	Standard
	Ho	165	778.4	7.8				ug/L	10	Standard
	Tl	203	123.7	14.4	**0.0129**	0.002	13.5	ug/L	4	Standard
	Tl	205	116.7	9.9	**0.0161**	0.001	9.2	ug/L	2	Standard
	Pb	206	819.0	5.9	**0.0785**	0.008	10.4	ug/L	240	Standard
	Pb	207	658.0	2.8	**0.0655**	0.007	10.9	ug/L	198	Standard
	Pb	208	2733.4	3.4	**0.0682**	0.002	2.2	ug/L	883	Standard
	U	238	29463.0	4.0	**4.1728**	0.071	1.7	ug/L	3	Standard
>	**Bi**	209	536825.9	2.9				ug/L	523984	Standard

Sample ID: L1606098603
Report Date/Time: Wednesday, June 29, 2016 13:15:19
Page 1

Approved: July 04, 2016

Na	23	18.3	31.5	**216.6166**	73.255	33.8	mg/L	0	Standard
Mg	24	7246.7	2.5	**9.4794**	0.084	0.9	mg/L	37	Standard
K	39	393.3	11.0	**0.7727**	0.091	11.8	mg/L	38	Standard
Ca	43	1938.5	3.9	**122.4130**	4.002	3.3	mg/L	10	Standard
Fe	54	1213.7	6.6	**0.9123**	0.078	8.6	mg/L	179	Standard
Fe	57	1360.1	3.5	**3.6434**	0.167	4.6	mg/L	335	Standard
> **Sc-1**	45	59811.2	2.1				mg/L	61029	Standard
Cl	35	41831.0	3.4				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	8062.1	5.5				ug/L	1547	Standard
P	31	19340.7	4.7				ug/L	15286	Standard
S	34	2026.8	7.6				ug/L	1572	Standard
Sr	88	1283.4	13.4				ug/L	73	Standard
C	12	153.3	19.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	1258.4	10.3				mg/L	6	Standard
Ho-1	165	778.4	7.8				mg/L	10	Standard
Er	166	800.0	9.9				mg/L	13	Standard
I	127	76815.4	9.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		143.190	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.817	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098603
Report Date/Time: Wednesday, June 29, 2016 13:15:19
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.214
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.451
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098603
Report Date/Time: Wednesday, June 29, 2016 13:15:19
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603PS WG574572-01
Sample Date/Time: Wednesday, June 29, 2016 13:16:14
Number of Replicates: 3
Autosampler Position: 316
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	107768.6	1.4				ug/L	74699	Standard
	Be	9	63154.7	2.1	**49.7293**	0.735	1.5	ug/L	8	Standard
	Al	27	62883924.9	0.4	**296.6156**	4.197	1.4	ug/L	1493	Standard
	Sc	45	60669.6	1.7				ug/L	61029	Standard
	Ti	47	431.3	11.1	**1.0313**	0.120	11.7	ug/L	35	Standard
	V	51	423481.5	0.7	**56.0150**	0.782	1.4	ug/L	1435	Standard
	Cr	52	394353.4	1.0	**52.3183**	0.613	1.2	ug/L	7635	Standard
	Cr	53	48385.5	1.4	**52.1545**	1.086	2.1	ug/L	363	Standard
	Mn	55	12676353.0	0.9	**1825.0939**	28.984	1.6	ug/L	747	Standard
	Co	59	406698.1	0.5	**61.7025**	0.702	1.1	ug/L	234	Standard
	Ni	60	105609.8	0.7	**56.4229**	0.793	1.4	ug/L	215	Standard
	Cu	65	109624.8	0.4	**58.9403**	0.355	0.6	ug/L	104	Standard
	Zn	66	80974.1	1.0	**73.8592**	1.163	1.6	ug/L	363	Standard
>	**Ge**	72	493944.8	0.7				ug/L	495142	Standard
	As	75	62963.2	0.4	**56.4099**	0.584	1.0	ug/L	-50	Standard
	Se	82	6528.2	1.4	**59.4919**	1.250	2.1	ug/L	14	Standard
	Se-1	77	4208.9	2.0	**57.1456**	1.224	2.1	ug/L	42	Standard
>	**Ga**	71	341.7	3.7				mg/L	42	Standard
	Rb	85	9437.9	2.3				ug/L	17	Standard
	Y	89	453637.5	2.0				ug/L	417098	Standard
>	**Rh**	103	333.3	4.8				ug/L	13	Standard
	Mo	98	1406.8	1.9	**0.5023**	0.008	1.6	ug/L	11	Standard
	Ag	107	336751.7	1.1	**49.7241**	0.838	1.7	ug/L	68	Standard
	Cd	111	110645.7	0.7	**50.9305**	0.669	1.3	mg/L	2	Standard
	Cd	114	268194.0	0.5	**50.3333**	0.438	0.9	ug/L	12	Standard
>	**In**	115	610323.2	0.7				ug/L	587441	Standard
	Sn	118	1538.4	2.3	**0.1588**	0.006	3.5	ug/L	497	Standard
	Sb	123	241381.6	1.1	**49.7617**	0.807	1.6	ug/L	78	Standard
	Ba	135	126394.4	1.4	**59.4274**	1.193	2.0	ug/L	24	Standard
	Ce	140	44334.8	0.8				ug/L	25	Standard
>	**Tb**	159	999714.8	1.0				ug/L	939701	Standard
	Ho	165	786.7	10.3				ug/L	10	Standard
	Tl	203	445077.8	1.6	**52.3048**	0.918	1.8	ug/L	4	Standard
	Tl	205	409895.5	2.2	**55.5017**	1.265	2.3	ug/L	2	Standard
	Pb	206	293321.0	0.4	**53.1218**	0.228	0.4	ug/L	240	Standard
	Pb	207	252054.7	0.9	**51.3779**	0.617	1.2	ug/L	198	Standard
	Pb	208	1026149.4	0.5	**52.9396**	0.195	0.4	ug/L	883	Standard
	U	238	420484.7	1.6	**59.1737**	0.984	1.7	ug/L	3	Standard
>	**Bi**	209	540895.8	0.3				ug/L	523984	Standard

Sample ID: L1606098603PS WG574572-01
Report Date/Time: Wednesday, June 29, 2016 13:18:31
Page 1

Approved: July 04, 2016

Na	23	16.7	45.8	**192.4294**	85.849	44.6	mg/L	0	Standard
Mg	24	7256.7	2.5	**9.3609**	0.349	3.7	mg/L	37	Standard
K	39	406.7	8.6	**0.7894**	0.064	8.1	mg/L	38	Standard
Ca	43	1853.4	4.6	**115.2974**	6.640	5.8	mg/L	10	Standard
Fe	54	1091.3	5.4	**0.7921**	0.036	4.5	mg/L	179	Standard
Fe	57	1315.1	6.5	**3.4198**	0.224	6.6	mg/L	335	Standard
> **Sc-1**	45	60669.6	1.7				mg/L	61029	Standard
Cl	35	43150.7	2.4				ug/L	20195	Standard
Kr	83	2.3	49.5				ug/L	4	Standard
Br	81	7932.1	5.6				ug/L	1547	Standard
P	31	18981.9	3.1				ug/L	15286	Standard
S	34	1910.1	1.6				ug/L	1572	Standard
Sr	88	1396.7	9.1				ug/L	73	Standard
C	12	166.7	22.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	1221.7	7.2				mg/L	6	Standard
Ho-1	165	786.7	10.3				mg/L	10	Standard
Er	166	800.0	7.6				mg/L	13	Standard
I	127	76722.5	5.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		144.271	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.758	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098603PS WG574572-01
Report Date/Time: Wednesday, June 29, 2016 13:18:31
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	103.895
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	103.227
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098603PS WG574572-01
Report Date/Time: Wednesday, June 29, 2016 13:18:31
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603SDL WG574572-02
Sample Date/Time: Wednesday, June 29, 2016 13:19:25
Number of Replicates: 3
Autosampler Position: 317
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	109627.0	2.2				ug/L	74699	Standard
	Be	9	55.0	32.8	**0.0501**	0.014	27.0	ug/L	8	Standard
	Al	27	13052757.9	2.1	**60.5215**	1.604	2.7	ug/L	1493	Standard
	Sc	45	62358.1	1.0				ug/L	61029	Standard
	Ti	47	123.0	14.7	**0.1926**	0.050	25.7	ug/L	35	Standard
	V	51	2451.6	1.0	**0.1057**	0.008	7.9	ug/L	1435	Standard
	Cr	52	10290.8	1.8	**0.2648**	0.008	2.8	ug/L	7635	Standard
	Cr	53	588.3	10.7	**0.2610**	0.063	24.3	ug/L	363	Standard
	Mn	55	2629853.6	1.6	**350.4744**	10.232	2.9	ug/L	747	Standard
	Co	59	4233.3	3.1	**0.5577**	0.018	3.3	ug/L	234	Standard
	Ni	60	2886.6	1.8	**1.3071**	0.037	2.8	ug/L	215	Standard
	Cu	65	3859.5	1.6	**1.8207**	0.021	1.2	ug/L	104	Standard
	Zn	66	7087.3	1.2	**5.5680**	0.153	2.7	ug/L	363	Standard
>	**Ge**	72	533577.2	1.8				ug/L	495142	Standard
	As	75	314.5	3.2	**0.3203**	0.010	3.2	ug/L	-50	Standard
	Se	82	28.2	16.0	**0.1004**	0.037	37.2	ug/L	14	Standard
	Se-1	77	51.7	9.9	**0.0349**	0.075	214.8	ug/L	42	Standard
>	**Ga**	71	153.3	21.0				mg/L	42	Standard
	Rb	85	1916.8	2.5				ug/L	17	Standard
	Y	89	457911.1	1.7				ug/L	417098	Standard
>	**Rh**	103	73.3	17.2				ug/L	13	Standard
	Mo	98	334.8	7.8	**0.1120**	0.009	8.5	ug/L	11	Standard
	Ag	107	142.0	14.1	**0.0075**	0.003	37.0	ug/L	68	Standard
	Cd	111	78.8	16.1	**0.0274**	0.006	20.4	mg/L	2	Standard
	Cd	114	171.7	29.4	**0.0293**	0.009	31.0	ug/L	12	Standard
>	**In**	115	626915.6	0.6				ug/L	587441	Standard
	Sn	118	618.3	15.8	**0.0044**	0.015	338.7	ug/L	497	Standard
	Sb	123	2031.6	21.7	**0.4032**	0.086	21.3	ug/L	78	Standard
	Ba	135	5054.8	0.4	**2.2983**	0.024	1.0	ug/L	24	Standard
	Ce	140	9302.9	0.9				ug/L	25	Standard
>	**Tb**	159	1014481.9	2.3				ug/L	939701	Standard
	Ho	165	175.0	2.9				ug/L	10	Standard
	Tl	203	125.0	17.4	**0.0120**	0.002	18.4	ug/L	4	Standard
	Tl	205	85.0	15.6	**0.0110**	0.002	15.7	ug/L	2	Standard
	Pb	206	501.3	9.2	**0.0141**	0.008	53.2	ug/L	240	Standard
	Pb	207	392.0	5.3	**0.0051**	0.005	98.5	ug/L	198	Standard
	Pb	208	1534.0	5.0	**0.0003**	0.005	1457.8	ug/L	883	Standard
	U	238	6033.2	1.4	**0.7925**	0.020	2.5	ug/L	3	Standard
>	**Bi**	209	576425.9	1.5				ug/L	523984	Standard

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:21:42
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**37.9583**	65.737	173.2	mg/L	0	Standard
Mg	24	1580.1	4.3	**1.9318**	0.105	5.4	mg/L	37	Standard
K	39	138.3	24.1	**0.1853**	0.073	39.3	mg/L	38	Standard
Ca	43	413.3	5.7	**23.0292**	1.446	6.3	mg/L	10	Standard
Fe	54	367.9	5.4	**0.1671**	0.019	11.4	mg/L	179	Standard
Fe	57	475.0	7.9	**0.5062**	0.132	26.1	mg/L	335	Standard
Sc-1	45	62358.1	1.0				mg/L	61029	Standard
Cl	35	34619.5	1.5				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	3197.0	3.6				ug/L	1547	Standard
P	31	18749.9	1.5				ug/L	15286	Standard
S	34	2010.1	3.0				ug/L	1572	Standard
Sr	88	305.0	20.5				ug/L	73	Standard
C	12	113.3	31.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	348.8	17.3				mg/L	6	Standard
Ho-1	165	175.0	2.9				mg/L	10	Standard
Er	166	163.3	52.8				mg/L	13	Standard
I	127	19115.4	2.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		146.759	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		107.762	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:21:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.720
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.008
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603SDL WG574572-02
Sample Date/Time: Wednesday, June 29, 2016 13:22:37
Number of Replicates: 3
Autosampler Position: 318
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	107882.7	0.8				ug/L	74699	Standard
	Be	9	30.0	44.1	**0.0312**	0.010	32.7	ug/L	8	Standard
	Al	27	2675117.5	0.8	**12.5917**	0.213	1.7	ug/L	1493	Standard
	Sc	45	65760.7	1.1				ug/L	61029	Standard
	Ti	47	48.3	10.6	**0.0072**	0.012	160.2	ug/L	35	Standard
	V	51	2266.6	5.1	**0.0781**	0.015	19.5	ug/L	1435	Standard
	Cr	52	10143.4	2.6	**0.2243**	0.020	8.9	ug/L	7635	Standard
	Cr	53	413.3	17.1	**0.0775**	0.065	83.8	ug/L	363	Standard
	Mn	55	539662.1	1.4	**70.5366**	0.569	0.8	ug/L	747	Standard
	Co	59	1015.0	2.6	**0.1030**	0.006	5.9	ug/L	234	Standard
	Ni	60	769.4	2.6	**0.2509**	0.010	4.1	ug/L	215	Standard
	Cu	65	948.0	4.2	**0.3607**	0.016	4.4	ug/L	104	Standard
	Zn	66	2642.9	1.7	**1.7518**	0.018	1.1	ug/L	363	Standard
>	**Ge**	72	543000.4	1.8				ug/L	495142	Standard
	As	75	12.5	225.4	**0.0698**	0.023	32.5	ug/L	-50	Standard
	Se	82	15.8	49.7	**-0.0063**	0.066	1051.8	ug/L	14	Standard
	Se-1	77	45.3	17.1	**-0.0563**	0.096	170.3	ug/L	42	Standard
>	**Ga**	71	85.0	15.6				mg/L	42	Standard
	Rb	85	428.3	11.2				ug/L	17	Standard
	Y	89	460200.9	1.8				ug/L	417098	Standard
>	**Rh**	103	35.0	24.7				ug/L	13	Standard
	Mo	98	83.6	11.1	**0.0226**	0.003	14.3	ug/L	11	Standard
	Ag	107	112.7	14.2	**0.0027**	0.002	80.0	ug/L	68	Standard
	Cd	111	20.9	22.0	**0.0011**	0.002	176.9	mg/L	2	Standard
	Cd	114	53.5	39.5	**0.0073**	0.004	50.6	ug/L	12	Standard
>	**In**	115	649578.1	0.2				ug/L	587441	Standard
	Sn	118	553.3	9.5	**-0.0090**	0.008	92.4	ug/L	497	Standard
	Sb	123	558.9	31.9	**0.1041**	0.035	33.4	ug/L	78	Standard
	Ba	135	1064.4	0.7	**0.4543**	0.003	0.7	ug/L	24	Standard
	Ce	140	1928.5	4.3				ug/L	25	Standard
>	**Tb**	159	1049822.6	1.3				ug/L	939701	Standard
	Ho	165	40.0	33.1				ug/L	10	Standard
	Tl	203	40.7	35.8	**0.0026**	0.002	59.7	ug/L	4	Standard
	Tl	205	45.0	38.5	**0.0057**	0.002	36.9	ug/L	2	Standard
	Pb	206	372.0	6.6	**-0.0102**	0.004	38.7	ug/L	240	Standard
	Pb	207	294.7	6.9	**-0.0156**	0.004	25.6	ug/L	198	Standard
	Pb	208	1159.4	2.3	**-0.0199**	0.001	7.3	ug/L	883	Standard
	U	238	1206.0	0.9	**0.1490**	0.002	1.2	ug/L	3	Standard
>	**Bi**	209	598311.7	0.5				ug/L	523984	Standard

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:24:54
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.6400**	30.545	173.2	mg/L	0	Standard
Mg	24	363.3	29.0	**0.3711**	0.131	35.3	mg/L	37	Standard
K	39	63.3	27.7	**0.0157**	0.035	226.7	mg/L	38	Standard
Ca	43	110.0	12.0	**3.9202**	0.707	18.0	mg/L	10	Standard
Fe	54	190.5	10.6	**0.0117**	0.017	142.3	mg/L	179	Standard
Fe	57	268.3	12.0	**-0.2271**	0.110	48.6	mg/L	335	Standard
> **Sc-1**	45	65760.7	1.1				mg/L	61029	Standard
Cl	35	34122.4	1.0				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	2286.8	8.5				ug/L	1547	Standard
P	31	19529.2	1.3				ug/L	15286	Standard
S	34	2301.8	4.6				ug/L	1572	Standard
Sr	88	130.0	32.9				ug/L	73	Standard
C	12	83.3	6.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	60.8	35.4				mg/L	6	Standard
Ho-1	165	40.0	33.1				mg/L	10	Standard
Er	166	53.3	28.6				mg/L	13	Standard
I	127	5874.5	8.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		144.424	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.666	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:24:54
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	110.578
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	114.185
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 13:25:50
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	96295.4	4.3				ug/L	74699	Standard
	Be	9	56739.5	4.7	**50.0144**	1.617	3.2	ug/L	8	Standard
	Al	27	8198811.7	1.9	**43.2979**	1.271	2.9	ug/L	1493	Standard
	Sc	45	60020.3	2.0				ug/L	61029	Standard
	Ti	47	38068.4	0.4	**97.0435**	0.523	0.5	ug/L	35	Standard
	V	51	407671.2	1.0	**52.0137**	0.145	0.3	ug/L	1435	Standard
	Cr	52	382711.8	0.9	**48.9253**	0.135	0.3	ug/L	7635	Standard
	Cr	53	47653.2	2.5	**49.5394**	1.027	2.1	ug/L	363	Standard
	Mn	55	366963.7	3.6	**50.8236**	1.362	2.7	ug/L	747	Standard
	Co	59	380449.1	2.9	**55.6813**	1.097	2.0	ug/L	234	Standard
	Ni	60	94585.7	1.1	**48.7401**	0.435	0.9	ug/L	215	Standard
	Cu	65	94763.8	1.5	**49.1432**	0.553	1.1	ug/L	104	Standard
	Zn	66	56752.8	2.3	**49.7949**	0.707	1.4	ug/L	363	Standard
>	Ge	72	511914.0	0.9				ug/L	495142	Standard
	As	75	57985.2	1.9	**50.1285**	0.568	1.1	ug/L	-50	Standard
	Se	82	5989.4	2.5	**52.6418**	0.964	1.8	ug/L	14	Standard
	Se-1	77	3921.2	2.5	**51.3010**	0.802	1.6	ug/L	42	Standard
>	Ga	71	220.0	19.4				mg/L	42	Standard
	Rb	85	765.0	2.0				ug/L	17	Standard
	Y	89	446724.4	3.3				ug/L	417098	Standard
>	Rh	103	20.0	43.3				ug/L	13	Standard
	Mo	98	280984.3	2.1	**102.0498**	2.098	2.1	ug/L	11	Standard
	Ag	107	335763.8	0.9	**49.8677**	1.010	2.0	ug/L	68	Standard
	Cd	111	105816.2	1.8	**48.9835**	0.569	1.2	mg/L	2	Standard
	Cd	114	257781.1	1.6	**48.6563**	0.589	1.2	ug/L	12	Standard
>	In	115	606960.1	2.8				ug/L	587441	Standard
	Sn	118	295693.2	2.0	**48.9061**	0.749	1.5	ug/L	497	Standard
	Sb	123	236523.1	0.9	**49.0438**	0.901	1.8	ug/L	78	Standard
	Ba	135	99992.2	1.5	**47.2781**	0.609	1.3	ug/L	24	Standard
	Ce	140	91.7	6.3				ug/L	25	Standard
>	Tb	159	985978.7	1.6				ug/L	939701	Standard
	Ho	165	11.7	89.2				ug/L	10	Standard
	Tl	203	433708.5	1.0	**49.3829**	0.257	0.5	ug/L	4	Standard
	Tl	205	391770.8	1.1	**51.3967**	0.163	0.3	ug/L	2	Standard
	Pb	206	284957.7	0.7	**50.0002**	0.477	1.0	ug/L	240	Standard
	Pb	207	255175.6	1.7	**50.3910**	0.211	0.4	ug/L	198	Standard
	Pb	208	1010912.5	1.9	**50.5251**	0.466	0.9	ug/L	883	Standard
	U	238	378758.8	0.6	**51.6451**	0.416	0.8	ug/L	3	Standard
>	Bi	209	558275.4	1.4				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 13:28:07
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4168.9	2.3	**5.4065**	0.021	0.4	mg/L	37	Standard
K	39	1983.5	2.5	**4.3432**	0.044	1.0	mg/L	38	Standard
Ca	43	106.7	26.7	**4.3258**	1.829	42.3	mg/L	10	Standard
Fe	54	6165.0	5.0	**5.1771**	0.265	5.1	mg/L	179	Standard
Fe	57	1720.1	9.9	**4.8644**	0.466	9.6	mg/L	335	Standard
Sc-1	45	60020.3	2.0				mg/L	61029	Standard
Cl	35	28213.0	5.7				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	1853.5	23.0				ug/L	1547	Standard
P	31	20223.5	3.5				ug/L	15286	Standard
S	34	2575.2	6.1				ug/L	1572	Standard
Sr	88	78.3	26.6				ug/L	73	Standard
C	12	100.0	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	9.2	116.5				mg/L	6	Standard
Ho-1	165	11.7	89.2				mg/L	10	Standard
Er	166	16.7	91.7				mg/L	13	Standard
I	127	2436.9	25.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	100.029		
Al	27	86.596		
Sc	45			
Ti	47	97.044		
V	51	104.027		
Cr	52	97.851		
Cr	53			
Mn	55	101.647		
Co	59	111.363		
Ni	60	97.480		
Cu	65	98.286		
Zn	66	99.590		
Ge	72		103.387	
As	75	100.257		
Se	82	105.284		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 13:28:07
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	102.050		
Ag	107	99.735		
Cd	111	97.967		
Cd	114			
> In	115		103.323	
Sn	118	97.812		
Sb	123	98.088		
Ba	135	94.556		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	98.766		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.050		
U	238	103.290		
> Bi	209		106.544	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 13:28:07
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 13:29:01
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94691.7	1.8				ug/L	74699	Standard
	Be	9	76.7	106.2	**0.0755**	0.071	94.2	ug/L	8	Standard
	Al	27	36843.1	136.1	**0.1814**	0.263	145.0	ug/L	1493	Standard
	Sc	45	60632.8	2.3				ug/L	61029	Standard
	Ti	47	43.0	36.1	**-0.0010**	0.038	3637.8	ug/L	35	Standard
	V	51	1519.7	11.1	**-0.0035**	0.019	532.9	ug/L	1435	Standard
	Cr	52	7045.6	5.2	**-0.1157**	0.044	37.8	ug/L	7635	Standard
	Cr	53	391.7	11.6	**0.0747**	0.044	59.2	ug/L	363	Standard
	Mn	55	7226.9	125.7	**0.8342**	1.228	147.2	ug/L	747	Standard
	Co	59	517.7	90.4	**0.0371**	0.066	179.1	ug/L	234	Standard
	Ni	60	246.3	39.3	**0.0014**	0.048	3317.1	ug/L	215	Standard
	Cu	65	191.7	56.6	**-0.0057**	0.054	949.5	ug/L	104	Standard
	Zn	66	417.3	10.8	**-0.0898**	0.036	40.1	ug/L	363	Standard
>	**Ge**	72	518398.0	1.4				ug/L	495142	Standard
	As	75	-30.5	76.0	**0.0334**	0.020	59.9	ug/L	-50	Standard
	Se	82	13.4	12.4	**-0.0212**	0.016	76.0	ug/L	14	Standard
	Se-1	77	53.3	15.9	**0.0746**	0.106	141.7	ug/L	42	Standard
>	**Ga**	71	155.0	18.0				mg/L	42	Standard
	Rb	85	21.7	93.3				ug/L	17	Standard
	Y	89	436745.4	0.7				ug/L	417098	Standard
>	**Rh**	103	15.0	33.3				ug/L	13	Standard
	Mo	98	315.2	32.1	**0.1069**	0.036	33.6	ug/L	11	Standard
	Ag	107	222.7	75.6	**0.0196**	0.025	125.0	ug/L	68	Standard
	Cd	111	57.8	114.5	**0.0184**	0.030	163.6	mg/L	2	Standard
	Cd	114	159.0	130.1	**0.0274**	0.038	140.1	ug/L	12	Standard
>	**In**	115	616374.3	0.2				ug/L	587441	Standard
	Sn	118	845.0	17.1	**0.0432**	0.024	55.2	ug/L	497	Standard
	Sb	123	593.6	13.2	**0.1170**	0.016	13.9	ug/L	78	Standard
	Ba	135	102.7	102.4	**0.0317**	0.049	153.8	ug/L	24	Standard
	Ce	140	58.3	55.8				ug/L	25	Standard
>	**Tb**	159	995160.6	1.6				ug/L	939701	Standard
	Ho	165	15.0	100.0				ug/L	10	Standard
	Tl	203	219.0	114.2	**0.0226**	0.028	122.3	ug/L	4	Standard
	Tl	205	381.7	141.6	**0.0491**	0.069	140.6	ug/L	2	Standard
	Pb	206	388.0	38.6	**-0.0042**	0.025	597.1	ug/L	240	Standard
	Pb	207	336.0	35.9	**-0.0047**	0.023	485.4	ug/L	198	Standard
	Pb	208	2012.1	79.6	**0.0246**	0.078	315.7	ug/L	883	Standard
	U	238	395.3	145.1	**0.0482**	0.076	157.9	ug/L	3	Standard
>	**Bi**	209	567355.9	1.2				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 13:31:18
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	58.3	34.6	**0.0111**	0.028	248.5	mg/L	37	Standard
K	39	40.0	12.5	**-0.0251**	0.012	46.0	mg/L	38	Standard
Ca	43	31.7	9.1	**-0.5078**	0.213	41.9	mg/L	10	Standard
Fe	54	155.5	6.6	**-0.0056**	0.009	155.5	mg/L	179	Standard
Fe	57	230.0	7.8	**-0.2862**	0.078	27.3	mg/L	335	Standard
> **Sc-1**	45	60632.8	2.3				mg/L	61029	Standard
Cl	35	27178.4	0.6				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	1736.8	2.2				ug/L	1547	Standard
P	31	19502.5	1.8				ug/L	15286	Standard
S	34	2301.8	3.7				ug/L	1572	Standard
Sr	88	75.0	26.7				ug/L	73	Standard
C	12	86.7	17.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	12.7	93.2				mg/L	6	Standard
Ho-1	165	15.0	100.0				mg/L	10	Standard
Er	166	13.3	86.6				mg/L	13	Standard
I	127	1935.1	25.0				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		104.697	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 13:31:18
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.925
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.277
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Mn	55	

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 13:31:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098605
Sample Date/Time: Wednesday, June 29, 2016 13:32:14
Number of Replicates: 3
Autosampler Position: 319
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93251.1	1.7				ug/L	74699	Standard
	Be	9	326.7	10.4	**0.3052**	0.035	11.6	ug/L	8	Standard
	Al	27	19740200.2	0.7	**107.5996**	1.185	1.1	ug/L	1493	Standard
	Sc	45	60365.1	2.3				ug/L	61029	Standard
	Ti	47	450.3	3.0	**1.0603**	0.030	2.8	ug/L	35	Standard
	V	51	4644.9	8.0	**0.4097**	0.047	11.5	ug/L	1435	Standard
	Cr	52	12104.2	0.3	**0.5838**	0.010	1.7	ug/L	7635	Standard
	Cr	53	928.4	3.1	**0.6586**	0.026	3.9	ug/L	363	Standard
	Mn	55	98559.3	4.8	**13.7841**	0.650	4.7	ug/L	747	Standard
	Co	59	3296.0	4.6	**0.4539**	0.021	4.7	ug/L	234	Standard
	Ni	60	3163.0	1.8	**1.5389**	0.038	2.4	ug/L	215	Standard
	Cu	65	32310.7	0.8	**16.9827**	0.169	1.0	ug/L	104	Standard
	Zn	66	47105.4	0.5	**41.9893**	0.365	0.9	ug/L	363	Standard
>	**Ge**	72	503086.3	0.5				ug/L	495142	Standard
	As	75	294.8	15.7	**0.3186**	0.040	12.5	ug/L	-50	Standard
	Se	82	52.3	21.7	**0.3309**	0.101	30.6	ug/L	14	Standard
	Se-1	77	70.7	11.4	**0.3303**	0.114	34.4	ug/L	42	Standard
>	**Ga**	71	295.0	9.0				mg/L	42	Standard
	Rb	85	9109.4	3.7				ug/L	17	Standard
	Y	89	448266.6	2.2				ug/L	417098	Standard
>	**Rh**	103	83.3	28.4				ug/L	13	Standard
	Mo	98	1843.0	5.2	**0.6500**	0.034	5.3	ug/L	11	Standard
	Ag	107	226.3	74.5	**0.0200**	0.024	121.9	ug/L	68	Standard
	Cd	111	104.1	84.7	**0.0392**	0.040	101.3	mg/L	2	Standard
	Cd	114	254.7	74.3	**0.0449**	0.035	77.3	ug/L	12	Standard
>	**In**	115	619535.9	1.6				ug/L	587441	Standard
	Sn	118	1750.1	13.2	**0.1895**	0.038	20.1	ug/L	497	Standard
	Sb	123	790.9	17.9	**0.1564**	0.028	18.0	ug/L	78	Standard
	Ba	135	72683.9	1.3	**33.6664**	0.969	2.9	ug/L	24	Standard
	Ce	140	5139.2	3.4				ug/L	25	Standard
>	**Tb**	159	985815.7	2.2				ug/L	939701	Standard
	Ho	165	223.3	9.3				ug/L	10	Standard
	Tl	203	317.0	96.7	**0.0353**	0.035	100.6	ug/L	4	Standard
	Tl	205	193.3	47.9	**0.0263**	0.012	46.4	ug/L	2	Standard
	Pb	206	1382.4	8.3	**0.1795**	0.018	10.0	ug/L	240	Standard
	Pb	207	1152.7	10.0	**0.1653**	0.022	13.1	ug/L	198	Standard
	Pb	208	4722.3	7.7	**0.1699**	0.018	10.3	ug/L	883	Standard
	U	238	9861.9	3.5	**1.3841**	0.043	3.1	ug/L	3	Standard
>	**Bi**	209	540670.7	1.7				ug/L	523984	Standard

Sample ID: L1606098605
Report Date/Time: Wednesday, June 29, 2016 13:34:31
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	6986.6	1.5	**9.0573**	0.310	3.4	mg/L	37	Standard
K	39	308.3	17.9	**0.5733**	0.109	19.0	mg/L	38	Standard
Ca	43	606.7	4.1	**36.2577**	2.485	6.9	mg/L	10	Standard
Fe	54	348.6	15.4	**0.1605**	0.045	28.3	mg/L	179	Standard
Fe	57	585.0	8.4	**0.9378**	0.194	20.7	mg/L	335	Standard
> **Sc-1**	45	60365.1	2.3				mg/L	61029	Standard
Cl	35	31561.5	3.4				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	4964.2	11.3				ug/L	1547	Standard
P	31	20053.3	1.8				ug/L	15286	Standard
S	34	2155.2	4.8				ug/L	1572	Standard
Sr	88	343.3	18.3				ug/L	73	Standard
C	12	163.3	34.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	294.3	16.7				mg/L	6	Standard
Ho-1	165	223.3	9.3				mg/L	10	Standard
Er	166	120.0	16.7				mg/L	13	Standard
I	127	9371.2	3.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		124.836	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.604	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098605
Report Date/Time: Wednesday, June 29, 2016 13:34:31
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		105.463
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.184
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098607
Sample Date/Time: Wednesday, June 29, 2016 13:35:26
Number of Replicates: 3
Autosampler Position: 320
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96566.8	1.9				ug/L	74699	Standard
	Be	9	85.0	29.4	**0.0821**	0.021	25.7	ug/L	8	Standard
	Al	27	20333697.9	3.0	**107.0230**	2.931	2.7	ug/L	1493	Standard
	Sc	45	62230.9	2.2				ug/L	61029	Standard
	Ti	47	287.7	7.2	**0.6119**	0.056	9.2	ug/L	35	Standard
	V	51	3941.3	2.6	**0.3001**	0.017	5.7	ug/L	1435	Standard
	Cr	52	11512.0	2.6	**0.4518**	0.053	11.8	ug/L	7635	Standard
	Cr	53	838.4	4.6	**0.5317**	0.044	8.4	ug/L	363	Standard
	Mn	55	502202.8	0.8	**68.3542**	0.721	1.1	ug/L	747	Standard
	Co	59	4570.7	1.4	**0.6201**	0.010	1.6	ug/L	234	Standard
	Ni	60	3293.0	0.7	**1.5464**	0.014	0.9	ug/L	215	Standard
	Cu	65	6694.1	1.0	**3.3123**	0.070	2.1	ug/L	104	Standard
	Zn	66	11579.4	0.2	**9.6129**	0.092	1.0	ug/L	363	Standard
>	**Ge**	72	521399.1	1.0				ug/L	495142	Standard
	As	75	65.2	31.2	**0.1148**	0.017	14.9	ug/L	-50	Standard
	Se	82	28.9	16.0	**0.1117**	0.041	36.5	ug/L	14	Standard
	Se-1	77	63.7	8.1	**0.2056**	0.068	33.2	ug/L	42	Standard
>	**Ga**	71	136.7	35.5				mg/L	42	Standard
	Rb	85	5054.2	2.4				ug/L	17	Standard
	Y	89	452894.0	2.8				ug/L	417098	Standard
>	**Rh**	103	165.0	28.9				ug/L	13	Standard
	Mo	98	1211.0	1.7	**0.4303**	0.008	1.9	ug/L	11	Standard
	Ag	107	106.3	11.0	**0.0028**	0.002	64.6	ug/L	68	Standard
	Cd	111	72.3	5.5	**0.0253**	0.002	7.2	mg/L	2	Standard
	Cd	114	189.4	18.5	**0.0333**	0.006	19.3	ug/L	12	Standard
>	**In**	115	612132.4	0.4				ug/L	587441	Standard
	Sn	118	1080.0	6.7	**0.0827**	0.012	15.0	ug/L	497	Standard
	Sb	123	567.1	13.2	**0.1124**	0.015	13.5	ug/L	78	Standard
	Ba	135	35457.1	0.1	**16.6089**	0.050	0.3	ug/L	24	Standard
	Ce	140	1880.1	4.7				ug/L	25	Standard
>	**Tb**	159	996852.4	0.7				ug/L	939701	Standard
	Ho	165	58.3	17.8				ug/L	10	Standard
	Tl	203	60.0	16.9	**0.0051**	0.001	23.4	ug/L	4	Standard
	Tl	205	60.0	16.7	**0.0081**	0.001	15.8	ug/L	2	Standard
	Pb	206	636.0	0.7	**0.0415**	0.001	3.4	ug/L	240	Standard
	Pb	207	529.0	3.2	**0.0355**	0.003	9.7	ug/L	198	Standard
	Pb	208	2291.4	2.0	**0.0416**	0.003	7.7	ug/L	883	Standard
	U	238	20180.8	2.1	**2.7702**	0.079	2.9	ug/L	3	Standard
>	**Bi**	209	553749.9	0.8				ug/L	523984	Standard

Sample ID: L1606098607
Report Date/Time: Wednesday, June 29, 2016 13:37:42
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**57.0696**	57.922	101.5	mg/L	0	Standard
Mg	24	5551.0	3.6	**6.9630**	0.244	3.5	mg/L	37	Standard
K	39	316.7	11.2	**0.5727**	0.081	14.1	mg/L	38	Standard
Ca	43	636.7	7.9	**36.9594**	3.698	10.0	mg/L	10	Standard
Fe	54	176.9	14.5	**0.0085**	0.018	212.8	mg/L	179	Standard
Fe	57	531.7	3.6	**0.6989**	0.084	12.0	mg/L	335	Standard
> **Sc-1**	45	62230.9	2.2				mg/L	61029	Standard
Cl	35	31386.5	1.7				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	3093.6	2.6				ug/L	1547	Standard
P	31	20912.8	5.9				ug/L	15286	Standard
S	34	2228.5	4.5				ug/L	1572	Standard
Sr	88	703.3	5.1				ug/L	73	Standard
C	12	126.7	40.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	102.2	30.1				mg/L	6	Standard
Ho-1	165	58.3	17.8				mg/L	10	Standard
Er	166	93.3	6.2				mg/L	13	Standard
I	127	29892.9	5.9				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		129.275	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		105.303	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606098607
Report Date/Time: Wednesday, June 29, 2016 13:37:42
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	104.203
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	105.681
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098609
Sample Date/Time: Wednesday, June 29, 2016 13:38:37
Number of Replicates: 3
Autosampler Position: 321
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96550.0	0.2				ug/L	74699	Standard
	Be	9	53.3	14.3	**0.0545**	0.007	12.5	ug/L	8	Standard
	Al	27	22878376.4	1.6	**120.4307**	1.945	1.6	ug/L	1493	Standard
	Sc	45	62031.7	0.8				ug/L	61029	Standard
	Ti	47	1850.4	1.6	**4.6226**	0.192	4.2	ug/L	35	Standard
	V	51	5461.8	0.6	**0.5052**	0.022	4.3	ug/L	1435	Standard
	Cr	52	13396.9	1.2	**0.7280**	0.058	8.0	ug/L	7635	Standard
	Cr	53	1881.8	4.3	**1.6444**	0.138	8.4	ug/L	363	Standard
	Mn	55	79733.3	1.7	**10.9451**	0.135	1.2	ug/L	747	Standard
	Co	59	2496.2	2.4	**0.3289**	0.007	2.0	ug/L	234	Standard
	Ni	60	4821.8	0.8	**2.3714**	0.068	2.9	ug/L	215	Standard
	Cu	65	4208.3	2.7	**2.0873**	0.088	4.2	ug/L	104	Standard
	Zn	66	20900.1	0.5	**18.0858**	0.567	3.1	ug/L	363	Standard
>	**Ge**	72	511251.4	2.6				ug/L	495142	Standard
	As	75	201.6	14.3	**0.2338**	0.023	9.7	ug/L	-50	Standard
	Se	82	54.1	5.4	**0.3394**	0.016	4.7	ug/L	14	Standard
	Se-1	77	103.0	11.2	**0.7418**	0.124	16.8	ug/L	42	Standard
>	**Ga**	71	510.0	8.0				mg/L	42	Standard
	Rb	85	13392.6	2.9				ug/L	17	Standard
	Y	89	437578.5	1.9				ug/L	417098	Standard
>	**Rh**	103	115.0	21.7				ug/L	13	Standard
	Mo	98	379.6	2.6	**0.1321**	0.004	2.9	ug/L	11	Standard
	Ag	107	107.7	0.5	**0.0031**	0.000	8.0	ug/L	68	Standard
	Cd	111	160.7	1.3	**0.0664**	0.001	1.0	mg/L	2	Standard
	Cd	114	395.9	2.3	**0.0726**	0.002	3.0	ug/L	12	Standard
>	**In**	115	607119.7	1.0				ug/L	587441	Standard
	Sn	118	1270.1	2.5	**0.1157**	0.007	5.8	ug/L	497	Standard
	Sb	123	941.4	6.3	**0.1909**	0.012	6.2	ug/L	78	Standard
	Ba	135	18081.1	0.3	**8.5321**	0.064	0.8	ug/L	24	Standard
	Ce	140	9564.7	2.3				ug/L	25	Standard
>	**Tb**	159	994932.9	1.3				ug/L	939701	Standard
	Ho	165	368.3	5.1				ug/L	10	Standard
	Tl	203	147.3	4.5	**0.0153**	0.001	4.3	ug/L	4	Standard
	Tl	205	150.0	15.3	**0.0202**	0.003	16.2	ug/L	2	Standard
	Pb	206	1807.1	1.2	**0.2512**	0.001	0.3	ug/L	240	Standard
	Pb	207	1489.4	1.7	**0.2291**	0.005	2.2	ug/L	198	Standard
	Pb	208	6035.1	2.4	**0.2326**	0.004	1.7	ug/L	883	Standard
	U	238	2872.6	2.3	**0.3933**	0.005	1.2	ug/L	3	Standard
>	**Bi**	209	549776.1	1.1				ug/L	523984	Standard

Sample ID: L1606098609
Report Date/Time: Wednesday, June 29, 2016 13:40:55
Page 1

Approved: July 04, 2016

Na	23	6.7	114.6	**75.5001**	86.744	114.9	mg/L	0	Standard
Mg	24	5847.8	3.3	**7.3629**	0.301	4.1	mg/L	37	Standard
K	39	155.0	19.6	**0.2229**	0.066	29.5	mg/L	38	Standard
Ca	43	803.4	3.4	**47.4095**	2.024	4.3	mg/L	10	Standard
Fe	54	459.0	5.6	**0.2447**	0.024	9.9	mg/L	179	Standard
Fe	57	691.7	11.1	**1.2400**	0.273	22.0	mg/L	335	Standard
> **Sc-1**	45	62031.7	0.8				mg/L	61029	Standard
Cl	35	35055.2	1.5				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	8335.6	0.3				ug/L	1547	Standard
P	31	19207.2	4.0				ug/L	15286	Standard
S	34	2091.8	7.3				ug/L	1572	Standard
Sr	88	440.0	15.9				ug/L	73	Standard
C	12	146.7	7.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	580.9	19.4				mg/L	6	Standard
Ho-1	165	368.3	5.1				mg/L	10	Standard
Er	166	330.0	13.9				mg/L	13	Standard
I	127	13649.5	1.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		129.252	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.253	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098609
Report Date/Time: Wednesday, June 29, 2016 13:40:55
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	103.350
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.922
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606098609
Report Date/Time: Wednesday, June 29, 2016 13:40:55
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098611
Sample Date/Time: Wednesday, June 29, 2016 13:41:50
Number of Replicates: 3
Autosampler Position: 322
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94928.2	1.7				ug/L	74699	Standard
	Be	9	611.7	2.1	**0.5543**	0.002	0.4	ug/L	8	Standard
	Al	27	35531075.2	1.8	**190.3104**	6.477	3.4	ug/L	1493	Standard
	Sc	45	58926.0	1.1				ug/L	61029	Standard
	Ti	47	1221.7	3.3	**3.0767**	0.071	2.3	ug/L	35	Standard
	V	51	7317.6	1.4	**0.7622**	0.017	2.2	ug/L	1435	Standard
	Cr	52	16086.9	1.7	**1.1219**	0.012	1.1	ug/L	7635	Standard
	Cr	53	9996.6	2.0	**10.3612**	0.096	0.9	ug/L	363	Standard
	Mn	55	583301.8	0.5	**82.6724**	0.812	1.0	ug/L	747	Standard
	Co	59	14960.7	0.8	**2.2024**	0.031	1.4	ug/L	234	Standard
	Ni	60	6882.6	4.3	**3.5089**	0.114	3.2	ug/L	215	Standard
	Cu	65	67223.1	1.5	**35.5977**	0.129	0.4	ug/L	104	Standard
	Zn	66	37202.6	1.0	**33.2122**	0.242	0.7	ug/L	363	Standard
>	**Ge**	72	500908.0	1.1				ug/L	495142	Standard
	As	75	1307.8	5.6	**1.2134**	0.054	4.4	ug/L	-50	Standard
	Se	82	362.7	2.0	**3.1282**	0.043	1.4	ug/L	14	Standard
	Se-1	77	555.0	5.1	**6.8873**	0.307	4.5	ug/L	42	Standard
>	**Ga**	71	435.0	11.9				mg/L	42	Standard
	Rb	85	18344.4	2.4				ug/L	17	Standard
	Y	89	436166.7	1.7				ug/L	417098	Standard
>	**Rh**	103	445.0	7.4				ug/L	13	Standard
	Mo	98	3054.9	0.8	**1.1522**	0.023	2.0	ug/L	11	Standard
	Ag	107	121.7	12.3	**0.0060**	0.002	41.3	ug/L	68	Standard
	Cd	111	41.7	14.4	**0.0123**	0.003	24.8	mg/L	2	Standard
	Cd	114	142.6	14.6	**0.0260**	0.004	14.4	ug/L	12	Standard
>	**In**	115	581588.5	1.2				ug/L	587441	Standard
	Sn	118	2201.8	2.3	**0.2861**	0.011	3.9	ug/L	497	Standard
	Sb	123	1347.5	5.7	**0.2874**	0.019	6.6	ug/L	78	Standard
	Ba	135	63997.2	1.5	**31.5730**	0.849	2.7	ug/L	24	Standard
	Ce	140	22939.0	3.1				ug/L	25	Standard
>	**Tb**	159	946929.2	0.6				ug/L	939701	Standard
	Ho	165	348.3	12.2				ug/L	10	Standard
	Tl	203	270.7	7.5	**0.0312**	0.002	6.9	ug/L	4	Standard
	Tl	205	231.7	2.5	**0.0327**	0.001	2.0	ug/L	2	Standard
	Pb	206	3400.0	1.8	**0.5672**	0.008	1.4	ug/L	240	Standard
	Pb	207	2794.6	0.7	**0.5207**	0.008	1.5	ug/L	198	Standard
	Pb	208	11397.0	1.2	**0.5356**	0.008	1.6	ug/L	883	Standard
	U	238	7265.4	1.1	**1.0543**	0.007	0.7	ug/L	3	Standard
>	**Bi**	209	522379.2	0.9				ug/L	523984	Standard

Sample ID: L1606098611
Report Date/Time: Wednesday, June 29, 2016 13:44:07
Page 1

Approved: July 04, 2016

Na	23	10.0	50.0	**119.8009**	60.997	50.9	mg/L	0	Standard
Mg	24	10421.9	2.0	**13.8711**	0.425	3.1	mg/L	37	Standard
K	39	453.3	5.1	**0.9237**	0.053	5.7	mg/L	38	Standard
Ca	43	1046.7	6.1	**65.9862**	4.937	7.5	mg/L	10	Standard
Fe	54	728.6	2.4	**0.5015**	0.016	3.2	mg/L	179	Standard
Fe	57	860.0	12.2	**1.9559**	0.399	20.4	mg/L	335	Standard
> **Sc-1**	45	58926.0	1.1				mg/L	61029	Standard
Cl	35	32129.4	1.2				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	75514.4	2.1				ug/L	1547	Standard
P	31	20090.0	4.4				ug/L	15286	Standard
S	34	2098.5	7.9				ug/L	1572	Standard
Sr	88	1583.4	4.7				ug/L	73	Standard
C	12	130.0	7.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	561.0	7.3				mg/L	6	Standard
Ho-1	165	348.3	12.2				mg/L	10	Standard
Er	166	256.7	6.0				mg/L	13	Standard
I	127	84570.2	9.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		127.081	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.164	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098611
Report Date/Time: Wednesday, June 29, 2016 13:44:07
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	99.004
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	99.694
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098613
Sample Date/Time: Wednesday, June 29, 2016 13:45:02
Number of Replicates: 3
Autosampler Position: 323
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101811.1	2.5				ug/L	74699	Standard
	Be	9	780.0	6.8	**0.6573**	0.030	4.5	ug/L	8	Standard
	Al	27	69460342.3	2.9	**346.7492**	1.873	0.5	ug/L	1493	Standard
	Sc	45	66852.2	1.8				ug/L	61029	Standard
	Ti	47	16395.2	2.4	**40.4139**	0.750	1.9	ug/L	35	Standard
	V	51	44929.4	0.9	**5.3782**	0.102	1.9	ug/L	1435	Standard
	Cr	52	38038.0	1.4	**3.7847**	0.045	1.2	ug/L	7635	Standard
	Cr	53	5357.6	0.6	**5.1013**	0.082	1.6	ug/L	363	Standard
	Mn	55	1205027.8	2.1	**162.0424**	5.610	3.5	ug/L	747	Standard
	Co	59	15267.0	2.9	**2.1291**	0.080	3.8	ug/L	234	Standard
	Ni	60	13379.6	2.6	**6.5697**	0.065	1.0	ug/L	215	Standard
	Cu	65	74766.0	1.3	**37.5331**	0.708	1.9	ug/L	104	Standard
	Zn	66	182914.0	1.7	**156.4214**	2.327	1.5	ug/L	363	Standard
>	**Ge**	72	528583.9	2.0				ug/L	495142	Standard
	As	75	2539.4	2.2	**2.1832**	0.013	0.6	ug/L	-50	Standard
	Se	82	108.7	5.6	**0.7905**	0.071	9.0	ug/L	14	Standard
	Se-1	77	150.3	9.2	**1.3050**	0.161	12.3	ug/L	42	Standard
>	**Ga**	71	5017.5	0.9				mg/L	42	Standard
	Rb	85	95755.1	2.6				ug/L	17	Standard
	Y	89	484844.4	1.2				ug/L	417098	Standard
>	**Rh**	103	451.7	17.2				ug/L	13	Standard
	Mo	98	2711.0	1.7	**0.9675**	0.010	1.1	ug/L	11	Standard
	Ag	107	547.0	8.0	**0.0673**	0.005	7.7	ug/L	68	Standard
	Cd	111	100.9	10.4	**0.0383**	0.005	14.0	mg/L	2	Standard
	Cd	114	263.2	6.1	**0.0470**	0.003	5.8	ug/L	12	Standard
>	**In**	115	613966.8	2.2				ug/L	587441	Standard
	Sn	118	1991.8	3.6	**0.2317**	0.015	6.6	ug/L	497	Standard
	Sb	123	971.4	9.0	**0.1947**	0.014	7.1	ug/L	78	Standard
	Ba	135	97214.2	0.8	**45.4386**	0.665	1.5	ug/L	24	Standard
	Ce	140	163522.7	0.6				ug/L	25	Standard
>	**Tb**	159	1011575.1	2.5				ug/L	939701	Standard
	Ho	165	4682.4	4.2				ug/L	10	Standard
	Tl	203	573.7	3.0	**0.0642**	0.003	4.6	ug/L	4	Standard
	Tl	205	525.0	7.2	**0.0698**	0.005	7.8	ug/L	2	Standard
	Pb	206	15938.7	1.8	**2.7562**	0.018	0.6	ug/L	240	Standard
	Pb	207	13053.6	0.6	**2.5367**	0.043	1.7	ug/L	198	Standard
	Pb	208	53401.0	1.9	**2.6245**	0.013	0.5	ug/L	883	Standard
	U	238	38818.0	2.5	**5.3391**	0.057	1.1	ug/L	3	Standard
>	**Bi**	209	553002.2	2.3				ug/L	523984	Standard

Sample ID: L1606098613
Report Date/Time: Wednesday, June 29, 2016 13:47:19
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**88.0394**	31.556	35.8	mg/L	0	Standard
Mg	24	10193.4	4.3	**11.9448**	0.378	3.2	mg/L	37	Standard
K	39	798.4	2.0	**1.4974**	0.060	4.0	mg/L	38	Standard
Ca	43	1498.4	4.4	**83.9112**	4.508	5.4	mg/L	10	Standard
Fe	54	4251.3	3.0	**3.1535**	0.141	4.5	mg/L	179	Standard
Fe	57	1956.8	0.6	**4.9983**	0.081	1.6	mg/L	335	Standard
Sc-1	45	66852.2	1.8				mg/L	61029	Standard
Cl	35	38555.7	1.3				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	16405.6	4.1				ug/L	1547	Standard
P	31	22243.0	1.3				ug/L	15286	Standard
S	34	1961.8	3.3				ug/L	1572	Standard
Sr	88	2096.8	8.9				ug/L	73	Standard
C	12	170.0	21.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	107.9				mg/L	3	Standard
Dy	164	7687.7	5.1				mg/L	6	Standard
Ho-1	165	4682.4	4.2				mg/L	10	Standard
Er	166	4197.2	1.1				mg/L	13	Standard
I	127	34873.4	0.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		136.295	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		106.754	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098613
Report Date/Time: Wednesday, June 29, 2016 13:47:19
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.515
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.538
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098613
Report Date/Time: Wednesday, June 29, 2016 13:47:19
Page 3

Approved: July 04, 2016

Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098615
Sample Date/Time: Wednesday, June 29, 2016 13:48:13
Number of Replicates: 3
Autosampler Position: 324
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	104439.3	4.8				ug/L	74699	Standard
	Be	9	536.7	12.4	**0.4427**	0.037	8.4	ug/L	8	Standard
	Al	27	77862989.2	8.3	**378.5121**	14.291	3.8	ug/L	1493	Standard
	Sc	45	58982.2	11.3				ug/L	61029	Standard
	Ti	47	2461.2	11.7	**6.5990**	0.434	6.6	ug/L	35	Standard
	V	51	4945.2	8.0	**0.4817**	0.017	3.5	ug/L	1435	Standard
	Cr	52	13134.4	7.3	**0.8107**	0.065	8.0	ug/L	7635	Standard
	Cr	53	1810.1	8.3	**1.6977**	0.117	6.9	ug/L	363	Standard
	Mn	55	1145096.1	8.2	**169.9719**	4.564	2.7	ug/L	747	Standard
	Co	59	4019.2	11.3	**0.5913**	0.032	5.4	ug/L	234	Standard
	Ni	60	7177.1	13.9	**3.8310**	0.314	8.2	ug/L	215	Standard
	Cu	65	41403.8	8.7	**22.8999**	0.588	2.6	ug/L	104	Standard
	Zn	66	43831.2	7.7	**41.0562**	0.750	1.8	ug/L	363	Standard
>	**Ge**	72	478398.7	6.6				ug/L	495142	Standard
	As	75	638.2	9.5	**0.6495**	0.045	6.9	ug/L	-50	Standard
	Se	82	102.9	17.3	**0.8353**	0.172	20.6	ug/L	14	Standard
	Se-1	77	94.0	19.4	**0.7027**	0.185	26.4	ug/L	42	Standard
>	**Ga**	71	463.3	11.0				mg/L	42	Standard
	Rb	85	23055.9	7.7				ug/L	17	Standard
	Y	89	432649.5	7.6				ug/L	417098	Standard
>	**Rh**	103	1188.4	10.3				ug/L	13	Standard
	Mo	98	1350.9	6.9	**0.5044**	0.013	2.5	ug/L	11	Standard
	Ag	107	139.0	6.4	**0.0086**	0.002	19.3	ug/L	68	Standard
	Cd	111	44.6	6.8	**0.0137**	0.002	16.4	mg/L	2	Standard
	Cd	114	129.2	31.8	**0.0237**	0.010	43.4	ug/L	12	Standard
>	**In**	115	583863.5	7.1				ug/L	587441	Standard
	Sn	118	2513.5	5.1	**0.3387**	0.012	3.5	ug/L	497	Standard
	Sb	123	705.1	6.8	**0.1478**	0.002	1.6	ug/L	78	Standard
	Ba	135	26452.2	7.2	**12.9866**	0.138	1.1	ug/L	24	Standard
	Ce	140	27398.2	9.1				ug/L	25	Standard
>	**Tb**	159	958605.6	6.1				ug/L	939701	Standard
	Ho	165	645.0	4.0				ug/L	10	Standard
	Tl	203	134.0	9.5	**0.0146**	0.001	7.1	ug/L	4	Standard
	Tl	205	118.3	6.5	**0.0169**	0.002	13.2	ug/L	2	Standard
	Pb	206	2508.9	7.1	**0.4009**	0.002	0.4	ug/L	240	Standard
	Pb	207	1976.5	6.7	**0.3487**	0.007	2.0	ug/L	198	Standard
	Pb	208	8160.9	6.7	**0.3634**	0.004	1.2	ug/L	883	Standard
	U	238	17292.6	7.2	**2.5205**	0.006	0.2	ug/L	3	Standard
>	**Bi**	209	521380.1	7.3				ug/L	523984	Standard

Sample ID: L1606098615
Report Date/Time: Wednesday, June 29, 2016 13:50:30
Page 1

Approved: July 04, 2016

Na	23	15.0	57.7	**176.9361**	92.757	52.4	mg/L	0	Standard
Mg	24	28440.1	8.6	**38.0068**	1.323	3.5	mg/L	37	Standard
K	39	673.3	1.9	**1.4418**	0.213	14.8	mg/L	38	Standard
Ca	43	1596.8	5.0	**102.4296**	8.312	8.1	mg/L	10	Standard
Fe	54	1077.2	5.0	**0.8118**	0.070	8.7	mg/L	179	Standard
Fe	57	1203.4	10.1	**3.1714**	0.327	10.3	mg/L	335	Standard
> **Sc-1**	45	58982.2	11.3				mg/L	61029	Standard
Cl	35	41482.2	7.1				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	15648.1	7.3				ug/L	1547	Standard
P	31	20014.9	6.8				ug/L	15286	Standard
S	34	1913.5	9.5				ug/L	1572	Standard
Sr	88	5429.3	11.3				ug/L	73	Standard
C	12	203.3	27.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	1127.2	20.3				mg/L	6	Standard
Ho-1	165	645.0	4.0				mg/L	10	Standard
Er	166	546.7	6.4				mg/L	13	Standard
I	127	112969.0	6.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		139.814	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.618	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098615
Report Date/Time: Wednesday, June 29, 2016 13:50:30
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.391
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.503
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098615
Report Date/Time: Wednesday, June 29, 2016 13:50:30
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098617
Sample Date/Time: Wednesday, June 29, 2016 13:51:25
Number of Replicates: 3
Autosampler Position: 325
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95101.0	2.9				ug/L	74699	Standard
	Be	9	61.7	9.4	**0.0626**	0.004	5.8	ug/L	8	Standard
	Al	27	16761639.5	0.8	**89.6118**	1.975	2.2	ug/L	1493	Standard
	Sc	45	53828.8	0.6				ug/L	61029	Standard
	Ti	47	756.4	12.3	**2.0144**	0.253	12.6	ug/L	35	Standard
	V	51	3040.4	3.6	**0.2328**	0.013	5.5	ug/L	1435	Standard
	Cr	52	11892.0	0.9	**0.6842**	0.022	3.3	ug/L	7635	Standard
	Cr	53	1066.7	2.7	**0.8986**	0.036	4.1	ug/L	363	Standard
	Mn	55	57447.7	1.9	**8.6332**	0.089	1.0	ug/L	747	Standard
	Co	59	1276.7	2.2	**0.1685**	0.005	2.7	ug/L	234	Standard
	Ni	60	2114.5	5.4	**1.0778**	0.056	5.2	ug/L	215	Standard
	Cu	65	5578.0	1.5	**3.0859**	0.022	0.7	ug/L	104	Standard
	Zn	66	8715.8	0.2	**8.0380**	0.061	0.8	ug/L	363	Standard
>	**Ge**	72	465226.7	0.9				ug/L	495142	Standard
	As	75	386.7	5.7	**0.4269**	0.018	4.2	ug/L	-50	Standard
	Se	82	42.7	8.0	**0.2761**	0.036	13.2	ug/L	14	Standard
	Se-1	77	67.0	16.1	**0.3532**	0.150	42.5	ug/L	42	Standard
>	**Ga**	71	183.3	18.2				mg/L	42	Standard
	Rb	85	8882.6	2.7				ug/L	17	Standard
	Y	89	409246.0	2.8				ug/L	417098	Standard
>	**Rh**	103	295.0	13.6				ug/L	13	Standard
	Mo	98	1514.9	3.6	**0.5995**	0.027	4.5	ug/L	11	Standard
	Ag	107	86.3	3.5	**0.0012**	0.001	48.9	ug/L	68	Standard
	Cd	111	45.4	6.6	**0.0152**	0.002	11.0	mg/L	2	Standard
	Cd	114	119.3	31.2	**0.0227**	0.008	33.9	ug/L	12	Standard
>	**In**	115	551838.0	2.1				ug/L	587441	Standard
	Sn	118	896.7	9.3	**0.0688**	0.017	24.4	ug/L	497	Standard
	Sb	123	476.3	5.5	**0.1045**	0.008	7.8	ug/L	78	Standard
	Ba	135	51101.0	1.7	**26.5634**	0.162	0.6	ug/L	24	Standard
	Ce	140	4162.2	9.3				ug/L	25	Standard
>	**Tb**	159	892953.9	1.2				ug/L	939701	Standard
	Ho	165	111.7	20.2				ug/L	10	Standard
	Tl	203	71.0	9.8	**0.0072**	0.001	12.5	ug/L	4	Standard
	Tl	205	50.0	10.0	**0.0075**	0.001	10.0	ug/L	2	Standard
	Pb	206	834.7	3.9	**0.0918**	0.006	7.0	ug/L	240	Standard
	Pb	207	668.0	3.0	**0.0768**	0.004	5.0	ug/L	198	Standard
	Pb	208	2813.8	1.3	**0.0825**	0.002	2.7	ug/L	883	Standard
	U	238	2276.2	2.7	**0.3405**	0.008	2.4	ug/L	3	Standard
>	**Bi**	209	502274.9	0.4				ug/L	523984	Standard

Sample ID: L1606098617
Report Date/Time: Wednesday, June 29, 2016 13:53:42
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**21.6273**	37.451	173.2	mg/L	0	Standard
Mg	24	11037.4	4.7	**16.0863**	0.689	4.3	mg/L	37	Standard
K	39	470.0	10.1	**1.0642**	0.126	11.9	mg/L	38	Standard
Ca	43	383.3	8.5	**24.9281**	2.309	9.3	mg/L	10	Standard
Fe	54	307.3	17.1	**0.1571**	0.051	32.4	mg/L	179	Standard
Fe	57	490.0	5.4	**0.8143**	0.108	13.3	mg/L	335	Standard
> **Sc-1**	45	53828.8	0.6				mg/L	61029	Standard
Cl	35	29673.0	1.2				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	4564.0	9.2				ug/L	1547	Standard
P	31	17653.6	5.5				ug/L	15286	Standard
S	34	1876.8	5.1				ug/L	1572	Standard
Sr	88	1206.7	6.6				ug/L	73	Standard
C	12	186.7	18.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	173.0	22.9				mg/L	6	Standard
Ho-1	165	111.7	20.2				mg/L	10	Standard
Er	166	76.7	30.1				mg/L	13	Standard
I	127	8810.9	16.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		127.313	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.958	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098617
Report Date/Time: Wednesday, June 29, 2016 13:53:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		93.939
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.857
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098619
Sample Date/Time: Wednesday, June 29, 2016 13:54:36
Number of Replicates: 3
Autosampler Position: 326
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96013.3	1.6				ug/L	74699	Standard
	Be	9	225.0	13.5	**0.2062**	0.024	11.6	ug/L	8	Standard
	Al	27	3833929.3	1.4	**20.2847**	0.231	1.1	ug/L	1493	Standard
	Sc	45	56757.9	4.8				ug/L	61029	Standard
	Ti	47	657.0	3.0	**1.6551**	0.023	1.4	ug/L	35	Standard
	V	51	6468.0	2.1	**0.6763**	0.022	3.3	ug/L	1435	Standard
	Cr	52	12143.2	0.5	**0.6446**	0.043	6.7	ug/L	7635	Standard
	Cr	53	1533.4	8.3	**1.3595**	0.147	10.8	ug/L	363	Standard
	Mn	55	41319.9	1.1	**5.8936**	0.094	1.6	ug/L	747	Standard
	Co	59	1041.4	1.4	**0.1233**	0.005	3.8	ug/L	234	Standard
	Ni	60	1381.7	2.6	**0.6272**	0.001	0.2	ug/L	215	Standard
	Cu	65	22006.0	2.1	**11.9309**	0.100	0.8	ug/L	104	Standard
	Zn	66	12936.5	1.8	**11.6003**	0.101	0.9	ug/L	363	Standard
>	**Ge**	72	486496.6	2.4				ug/L	495142	Standard
	As	75	1429.3	4.7	**1.3578**	0.030	2.2	ug/L	-50	Standard
	Se	82	368.2	2.0	**3.2780**	0.104	3.2	ug/L	14	Standard
	Se-1	77	166.3	2.4	**1.6971**	0.106	6.2	ug/L	42	Standard
>	**Ga**	71	256.7	8.8				mg/L	42	Standard
	Rb	85	6678.1	5.5				ug/L	17	Standard
	Y	89	411615.1	2.5				ug/L	417098	Standard
>	**Rh**	103	45.0	40.1				ug/L	13	Standard
	Mo	98	2416.6	3.5	**0.9418**	0.055	5.8	ug/L	11	Standard
	Ag	107	106.0	14.6	**0.0040**	0.002	49.0	ug/L	68	Standard
	Cd	111	306.3	1.8	**0.1451**	0.002	1.4	mg/L	2	Standard
	Cd	114	658.6	9.3	**0.1319**	0.011	8.6	ug/L	12	Standard
>	**In**	115	562685.3	3.0				ug/L	587441	Standard
	Sn	118	1296.7	5.8	**0.1370**	0.009	6.5	ug/L	497	Standard
	Sb	123	1866.5	3.1	**0.4132**	0.007	1.8	ug/L	78	Standard
	Ba	135	49124.5	1.8	**25.0474**	0.324	1.3	ug/L	24	Standard
	Ce	140	4487.3	4.7				ug/L	25	Standard
>	**Tb**	159	930596.3	3.6				ug/L	939701	Standard
	Ho	165	203.3	14.2				ug/L	10	Standard
	Tl	203	145.3	15.0	**0.0163**	0.002	14.2	ug/L	4	Standard
	Tl	205	131.7	4.4	**0.0191**	0.000	1.8	ug/L	2	Standard
	Pb	206	1800.8	1.2	**0.2755**	0.007	2.5	ug/L	240	Standard
	Pb	207	1451.7	2.6	**0.2445**	0.002	0.8	ug/L	198	Standard
	Pb	208	5832.1	0.4	**0.2459**	0.009	3.7	ug/L	883	Standard
	U	238	34205.2	1.0	**5.1077**	0.154	3.0	ug/L	3	Standard
>	**Bi**	209	509656.0	3.1				ug/L	523984	Standard

Sample ID: L1606098619
Report Date/Time: Wednesday, June 29, 2016 13:56:53
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	92209.8	1.7	**128.1856**	8.381	6.5	mg/L	37	Standard
K	39	298.3	3.5	**0.5957**	0.033	5.6	mg/L	38	Standard
Ca	43	80.0	6.3	**2.9252**	0.551	18.8	mg/L	10	Standard
Fe	54	254.0	14.7	**0.0930**	0.030	32.4	mg/L	179	Standard
Fe	57	333.3	5.3	**0.1453**	0.073	50.0	mg/L	335	Standard
> **Sc-1**	45	56757.9	4.8				mg/L	61029	Standard
Cl	35	31152.7	1.5				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	53109.7	4.4				ug/L	1547	Standard
P	31	17403.3	2.0				ug/L	15286	Standard
S	34	1883.5	7.3				ug/L	1572	Standard
Sr	88	213.3	3.6				ug/L	73	Standard
C	12	256.7	12.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	288.4	16.8				mg/L	6	Standard
Ho-1	165	203.3	14.2				mg/L	10	Standard
Er	166	173.3	23.3				mg/L	13	Standard
I	127	470221.1	5.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		128.534	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.254	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098619
Report Date/Time: Wednesday, June 29, 2016 13:56:53
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.786
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.265
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098625
Sample Date/Time: Wednesday, June 29, 2016 13:57:48
Number of Replicates: 3
Autosampler Position: 327
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	102613.0	0.8				ug/L	74699	Standard
	Be	9	78.3	9.8	**0.0724**	0.006	8.8	ug/L	8	Standard
	Al	27	33905192.3	1.5	**167.9500**	3.413	2.0	ug/L	1493	Standard
	Sc	45	63494.4	0.7				ug/L	61029	Standard
	Ti	47	6149.2	2.1	**15.7581**	0.447	2.8	ug/L	35	Standard
	V	51	12969.6	1.6	**1.4837**	0.027	1.8	ug/L	1435	Standard
	Cr	52	17892.2	0.6	**1.3372**	0.031	2.3	ug/L	7635	Standard
	Cr	53	3548.7	0.7	**3.4248**	0.009	0.3	ug/L	363	Standard
	Mn	55	3105877.1	1.7	**436.1197**	10.576	2.4	ug/L	747	Standard
	Co	59	6022.9	2.6	**0.8547**	0.030	3.5	ug/L	234	Standard
	Ni	60	6318.7	1.1	**3.1764**	0.053	1.7	ug/L	215	Standard
	Cu	65	5119.5	1.0	**2.5860**	0.011	0.4	ug/L	104	Standard
	Zn	66	19836.6	0.9	**17.3044**	0.276	1.6	ug/L	363	Standard
>	Ge	72	506357.7	0.8				ug/L	495142	Standard
	As	75	1090.3	6.9	**1.0117**	0.071	7.0	ug/L	-50	Standard
	Se	82	334.6	3.6	**2.8439**	0.125	4.4	ug/L	14	Standard
	Se-1	77	270.0	7.0	**2.9935**	0.272	9.1	ug/L	42	Standard
>	Ga	71	1363.4	8.0				mg/L	42	Standard
	Rb	85	42725.2	0.9				ug/L	17	Standard
	Y	89	453938.8	1.4				ug/L	417098	Standard
>	Rh	103	198.3	19.1				ug/L	13	Standard
	Mo	98	1348.4	1.8	**0.4789**	0.005	1.1	ug/L	11	Standard
	Ag	107	259.3	4.5	**0.0252**	0.002	7.6	ug/L	68	Standard
	Cd	111	351.0	2.3	**0.1529**	0.002	1.6	mg/L	2	Standard
	Cd	114	808.7	5.9	**0.1490**	0.010	6.6	ug/L	12	Standard
>	In	115	613191.6	0.9				ug/L	587441	Standard
	Sn	118	795.0	1.9	**0.0357**	0.003	9.4	ug/L	497	Standard
	Sb	123	319.9	2.9	**0.0614**	0.002	3.0	ug/L	78	Standard
	Ba	135	112678.5	1.7	**52.7237**	0.702	1.3	ug/L	24	Standard
	Ce	140	37260.8	2.0				ug/L	25	Standard
>	Tb	159	1001354.7	0.4				ug/L	939701	Standard
	Ho	165	1051.7	3.8				ug/L	10	Standard
	Tl	203	322.7	6.7	**0.0353**	0.003	7.6	ug/L	4	Standard
	Tl	205	283.3	17.7	**0.0376**	0.006	16.6	ug/L	2	Standard
	Pb	206	8100.8	1.1	**1.3649**	0.025	1.8	ug/L	240	Standard
	Pb	207	6526.4	1.1	**1.2323**	0.027	2.2	ug/L	198	Standard
	Pb	208	27115.2	0.7	**1.2953**	0.018	1.4	ug/L	883	Standard
	U	238	26591.0	2.4	**3.6542**	0.116	3.2	ug/L	3	Standard
>	Bi	209	553363.8	1.0				ug/L	523984	Standard

Sample ID: L1606098625
Report Date/Time: Wednesday, June 29, 2016 14:00:05
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**55.1859**	55.233	100.1	mg/L	0	Standard
Mg	24	8027.1	0.4	**9.8947**	0.087	0.9	mg/L	37	Standard
K	39	523.3	14.3	**0.9970**	0.152	15.2	mg/L	38	Standard
Ca	43	861.7	11.2	**49.8050**	6.018	12.1	mg/L	10	Standard
Fe	54	1108.4	4.4	**0.7648**	0.035	4.5	mg/L	179	Standard
Fe	57	975.0	10.7	**2.1112**	0.341	16.2	mg/L	335	Standard
> **Sc-1**	45	63494.4	0.7				mg/L	61029	Standard
Cl	35	33805.7	1.3				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	35418.0	3.4				ug/L	1547	Standard
P	31	20113.3	1.1				ug/L	15286	Standard
S	34	2053.5	4.6				ug/L	1572	Standard
Sr	88	896.7	5.2				ug/L	73	Standard
C	12	130.0	35.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	1595.5	8.0				mg/L	6	Standard
Ho-1	165	1051.7	3.8				mg/L	10	Standard
Er	166	1000.0	4.0				mg/L	13	Standard
I	127	43091.5	8.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		137.369	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.265	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098625
Report Date/Time: Wednesday, June 29, 2016 14:00:05
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.384
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.607
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098625
Report Date/Time: Wednesday, June 29, 2016 14:00:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098627
Sample Date/Time: Wednesday, June 29, 2016 14:00:59
Number of Replicates: 3
Autosampler Position: 328
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98891.7	1.0				ug/L	74699	Standard
	Be	9	638.3	5.3	**0.5552**	0.024	4.3	ug/L	8	Standard
	Al	27	63228292.3	2.4	**324.9422**	4.583	1.4	ug/L	1493	Standard
	Sc	45	59185.4	3.1				ug/L	61029	Standard
	Ti	47	12099.8	1.5	**31.9570**	0.495	1.6	ug/L	35	Standard
	V	51	31610.6	0.8	**4.0089**	0.089	2.2	ug/L	1435	Standard
	Cr	52	29501.7	2.0	**2.9745**	0.066	2.2	ug/L	7635	Standard
	Cr	53	4110.6	4.5	**4.1359**	0.142	3.4	ug/L	363	Standard
	Mn	55	939161.3	0.8	**135.3408**	0.783	0.6	ug/L	747	Standard
	Co	59	11927.7	1.5	**1.7769**	0.013	0.7	ug/L	234	Standard
	Ni	60	10174.7	1.2	**5.3344**	0.025	0.5	ug/L	215	Standard
	Cu	65	63138.6	0.6	**33.9717**	0.415	1.2	ug/L	104	Standard
	Zn	66	96624.1	0.5	**88.3979**	1.021	1.2	ug/L	363	Standard
>	**Ge**	72	492984.7	1.4				ug/L	495142	Standard
	As	75	1814.9	1.7	**1.6870**	0.014	0.9	ug/L	-50	Standard
	Se	82	95.5	0.5	**0.7363**	0.016	2.2	ug/L	14	Standard
	Se-1	77	118.3	17.1	**1.0027**	0.256	25.5	ug/L	42	Standard
>	**Ga**	71	3340.4	1.3				mg/L	42	Standard
	Rb	85	72434.4	1.9				ug/L	17	Standard
	Y	89	474043.5	0.9				ug/L	417098	Standard
>	**Rh**	103	386.7	16.4				ug/L	13	Standard
	Mo	98	2396.5	0.7	**0.9091**	0.018	1.9	ug/L	11	Standard
	Ag	107	338.3	4.7	**0.0399**	0.003	6.7	ug/L	68	Standard
	Cd	111	101.2	9.9	**0.0414**	0.005	12.6	mg/L	2	Standard
	Cd	114	230.2	21.2	**0.0435**	0.009	20.9	ug/L	12	Standard
>	**In**	115	577417.6	1.2				ug/L	587441	Standard
	Sn	118	1793.4	6.6	**0.2178**	0.023	10.5	ug/L	497	Standard
	Sb	123	600.3	4.8	**0.1267**	0.008	6.1	ug/L	78	Standard
	Ba	135	70114.5	1.5	**34.8357**	0.414	1.2	ug/L	24	Standard
	Ce	140	125576.7	1.8				ug/L	25	Standard
>	**Tb**	159	953741.7	1.0				ug/L	939701	Standard
	Ho	165	3755.5	3.8				ug/L	10	Standard
	Tl	203	452.7	12.3	**0.0539**	0.006	11.6	ug/L	4	Standard
	Tl	205	385.0	10.1	**0.0548**	0.005	9.0	ug/L	2	Standard
	Pb	206	12183.9	0.5	**2.2443**	0.020	0.9	ug/L	240	Standard
	Pb	207	9920.2	0.8	**2.0519**	0.018	0.9	ug/L	198	Standard
	Pb	208	42107.7	1.6	**2.2055**	0.026	1.2	ug/L	883	Standard
	U	238	36138.4	1.7	**5.3249**	0.032	0.6	ug/L	3	Standard
>	**Bi**	209	516181.2	1.2				ug/L	523984	Standard

Sample ID: L1606098627
Report Date/Time: Wednesday, June 29, 2016 14:03:16
Page 1

Approved: July 04, 2016

Na	23	8.3	91.7	**99.8498**	90.493	90.6	mg/L	0	Standard
Mg	24	9524.7	1.9	**12.6179**	0.315	2.5	mg/L	37	Standard
K	39	723.4	10.8	**1.5367**	0.198	12.9	mg/L	38	Standard
Ca	43	1308.4	8.3	**82.6186**	4.814	5.8	mg/L	10	Standard
Fe	54	3102.8	2.2	**2.5749**	0.034	1.3	mg/L	179	Standard
Fe	57	1665.1	3.0	**4.7616**	0.022	0.5	mg/L	335	Standard
> **Sc-1**	45	59185.4	3.1				mg/L	61029	Standard
Cl	35	38114.5	0.9				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	13185.7	5.6				ug/L	1547	Standard
P	31	20465.5	3.4				ug/L	15286	Standard
S	34	1960.1	7.7				ug/L	1572	Standard
Sr	88	1940.1	7.4				ug/L	73	Standard
C	12	216.7	23.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	3	Standard
Dy	164	5933.6	2.4				mg/L	6	Standard
Ho-1	165	3755.5	3.8				mg/L	10	Standard
Er	166	3430.4	0.6				mg/L	13	Standard
I	127	31850.1	1.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.387	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.564	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098627
Report Date/Time: Wednesday, June 29, 2016 14:03:16
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.294
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.511
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098627
Report Date/Time: Wednesday, June 29, 2016 14:03:16
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 14:04:12
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	95022.8	6.4				ug/L	74699	Standard
	Be	9	57413.6	2.5	**51.3575**	1.980	3.9	ug/L	8	Standard
	Al	27	7949888.2	1.3	**42.6051**	2.208	5.2	ug/L	1493	Standard
	Sc	45	58088.0	4.8				ug/L	61029	Standard
	Ti	47	37651.4	1.9	**100.2698**	3.249	3.2	ug/L	35	Standard
	V	51	404346.6	1.8	**53.8882**	0.844	1.6	ug/L	1435	Standard
	Cr	52	381698.1	1.1	**51.0223**	1.642	3.2	ug/L	7635	Standard
	Cr	53	47042.9	0.3	**51.1128**	1.842	3.6	ug/L	363	Standard
	Mn	55	357326.2	3.2	**51.6830**	0.417	0.8	ug/L	747	Standard
	Co	59	371699.6	2.5	**56.8155**	0.522	0.9	ug/L	234	Standard
	Ni	60	94541.7	1.4	**50.9011**	1.597	3.1	ug/L	215	Standard
	Cu	65	94983.8	1.5	**51.4773**	2.133	4.1	ug/L	104	Standard
	Zn	66	56900.0	0.7	**52.1829**	1.541	3.0	ug/L	363	Standard
>	Ge	72	490310.3	3.4				ug/L	495142	Standard
	As	75	57867.9	1.8	**52.2509**	0.921	1.8	ug/L	-50	Standard
	Se	82	5817.8	2.4	**53.4067**	0.847	1.6	ug/L	14	Standard
	Se-1	77	4012.5	1.4	**54.8814**	1.095	2.0	ug/L	42	Standard
>	Ga	71	80.0	6.3				mg/L	42	Standard
	Rb	85	831.7	8.5				ug/L	17	Standard
	Y	89	423000.5	2.4				ug/L	417098	Standard
>	Rh	103	26.7	57.3				ug/L	13	Standard
	Mo	98	279510.1	1.2	**103.0732**	1.481	1.4	ug/L	11	Standard
	Ag	107	337025.9	0.7	**50.8187**	0.653	1.3	ug/L	68	Standard
	Cd	111	104719.9	1.6	**49.2197**	0.492	1.0	mg/L	2	Standard
	Cd	114	254484.5	1.8	**48.7789**	1.357	2.8	ug/L	12	Standard
>	In	115	597695.8	1.6				ug/L	587441	Standard
	Sn	118	297589.5	3.3	**49.9660**	0.932	1.9	ug/L	497	Standard
	Sb	123	237800.4	0.6	**50.0611**	0.480	1.0	ug/L	78	Standard
	Ba	135	98785.4	1.3	**47.4247**	0.676	1.4	ug/L	24	Standard
	Ce	140	118.3	6.5				ug/L	25	Standard
>	Tb	159	962162.8	1.8				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	431721.7	1.4	**50.5906**	0.299	0.6	ug/L	4	Standard
	Tl	205	386667.7	1.8	**52.2055**	0.331	0.6	ug/L	2	Standard
	Pb	206	286495.7	0.6	**51.7458**	1.032	2.0	ug/L	240	Standard
	Pb	207	254256.8	1.2	**51.6830**	0.707	1.4	ug/L	198	Standard
	Pb	208	993731.8	1.4	**51.1193**	0.294	0.6	ug/L	883	Standard
	U	238	371822.3	1.0	**52.1805**	0.682	1.3	ug/L	3	Standard
>	Bi	209	542470.6	1.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 14:06:29
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**20.6892**	35.826	173.2	mg/L	0	Standard
Mg	24	4115.6	5.8	**5.5148**	0.066	1.2	mg/L	37	Standard
K	39	1971.8	1.5	**4.4711**	0.214	4.8	mg/L	38	Standard
Ca	43	71.7	16.1	**2.2502**	0.891	39.6	mg/L	10	Standard
Fe	54	5897.1	1.5	**5.1229**	0.272	5.3	mg/L	179	Standard
Fe	57	1798.4	2.6	**5.3592**	0.387	7.2	mg/L	335	Standard
Sc-1	45	58088.0	4.8				mg/L	61029	Standard
Cl	35	26443.8	4.7				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	1943.5	9.9				ug/L	1547	Standard
P	31	18503.0	1.3				ug/L	15286	Standard
S	34	2213.5	3.2				ug/L	1572	Standard
Sr	88	76.7	10.0				ug/L	73	Standard
C	12	120.0	43.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	2.9	217.0				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	3483.8	25.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	102.715		
Al	27	85.210		
Sc	45			
Ti	47	100.270		
V	51	107.776		
Cr	52	102.045		
Cr	53			
Mn	55	103.366		
Co	59	113.631		
Ni	60	101.802		
Cu	65	102.955		
Zn	66	104.366		
Ge	72		99.024	
As	75	104.502		
Se	82	106.813		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 14:06:29
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	103.073		
Ag	107	101.637		
Cd	111	98.439		
Cd	114			
In	115		101.746	
Sn	118	99.932		
Sb	123	100.122		
Ba	135	94.849		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.181		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.239		
U	238	104.361		
Bi	209		103.528	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 14:06:29
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 14:07:23
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87483.9	4.7				ug/L	74699	Standard
	Be	9	21.7	96.1	**0.0292**	0.021	73.0	ug/L	8	Standard
	Al	27	18230.8	94.8	**0.0950**	0.107	112.2	ug/L	1493	Standard
	Sc	45	56471.8	3.5				ug/L	61029	Standard
	Ti	47	36.0	46.7	**-0.0098**	0.050	510.8	ug/L	35	Standard
	V	51	1334.8	2.3	**-0.0125**	0.010	80.4	ug/L	1435	Standard
	Cr	52	6684.1	1.9	**-0.0897**	0.031	35.0	ug/L	7635	Standard
	Cr	53	393.3	6.4	**0.1115**	0.043	38.6	ug/L	363	Standard
	Mn	55	1346.4	57.0	**0.0546**	0.123	224.6	ug/L	747	Standard
	Co	59	278.3	17.3	**0.0065**	0.009	139.3	ug/L	234	Standard
	Ni	60	177.0	5.9	**-0.0261**	0.004	17.0	ug/L	215	Standard
	Cu	65	145.0	10.8	**-0.0230**	0.009	41.3	ug/L	104	Standard
	Zn	66	399.3	2.4	**-0.0757**	0.021	28.0	ug/L	363	Standard
>	**Ge**	72	478426.6	3.3				ug/L	495142	Standard
	As	75	-39.3	49.9	**0.0231**	0.019	81.6	ug/L	-50	Standard
	Se	82	16.5	11.7	**0.0181**	0.023	124.9	ug/L	14	Standard
	Se-1	77	38.7	26.9	**-0.0773**	0.131	170.1	ug/L	42	Standard
>	**Ga**	71	66.7	15.6				mg/L	42	Standard
	Rb	85	18.3	41.7				ug/L	17	Standard
	Y	89	416242.9	2.8				ug/L	417098	Standard
>	**Rh**	103	15.0	57.7				ug/L	13	Standard
	Mo	98	272.1	23.3	**0.0976**	0.024	24.4	ug/L	11	Standard
	Ag	107	106.0	6.8	**0.0036**	0.001	28.2	ug/L	68	Standard
	Cd	111	10.2	22.2	**-0.0030**	0.001	37.6	mg/L	2	Standard
	Cd	114	31.6	53.8	**0.0041**	0.003	80.7	ug/L	12	Standard
>	**In**	115	580179.3	0.7				ug/L	587441	Standard
	Sn	118	746.7	16.1	**0.0348**	0.022	62.5	ug/L	497	Standard
	Sb	123	368.4	24.4	**0.0758**	0.020	26.4	ug/L	78	Standard
	Ba	135	41.0	10.6	**0.0043**	0.002	51.4	ug/L	24	Standard
	Ce	140	38.3	61.6				ug/L	25	Standard
>	**Tb**	159	945404.9	0.5				ug/L	939701	Standard
	Ho	165	6.7	43.3				ug/L	10	Standard
	Tl	203	65.7	8.7	**0.0061**	0.001	8.8	ug/L	4	Standard
	Tl	205	95.0	45.9	**0.0133**	0.006	45.0	ug/L	2	Standard
	Pb	206	276.0	7.9	**-0.0202**	0.003	15.8	ug/L	240	Standard
	Pb	207	258.0	12.6	**-0.0163**	0.006	36.3	ug/L	198	Standard
	Pb	208	1134.3	12.6	**-0.0143**	0.008	56.7	ug/L	883	Standard
	U	238	91.0	75.8	**0.0087**	0.010	115.9	ug/L	3	Standard
>	**Bi**	209	530301.2	1.8				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 14:09:40
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	81.1	**0.0007**	0.054	7923.6	mg/L	37	Standard
K	39	40.0	45.1	**-0.0194**	0.039	202.6	mg/L	38	Standard
Ca	43	28.3	20.4	**-0.5841**	0.427	73.1	mg/L	10	Standard
Fe	54	129.2	39.5	**-0.0191**	0.049	258.2	mg/L	179	Standard
Fe	57	241.7	27.6	**-0.1803**	0.270	149.7	mg/L	335	Standard
> **Sc-1**	45	56471.8	3.5				mg/L	61029	Standard
Cl	35	26027.0	1.3				ug/L	20195	Standard
Kr	83	3.0	57.7				ug/L	4	Standard
Br	81	1783.4	6.5				ug/L	1547	Standard
P	31	18152.5	2.7				ug/L	15286	Standard
S	34	2168.5	1.2				ug/L	1572	Standard
Sr	88	56.7	27.0				ug/L	73	Standard
C	12	93.3	24.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	6.5	86.7				mg/L	6	Standard
Ho-1	165	6.7	43.3				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	2606.9	21.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.624	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 14:09:40
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	98.764	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	101.206	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 14:09:40
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098629
Sample Date/Time: Wednesday, June 29, 2016 14:10:36
Number of Replicates: 3
Autosampler Position: 329
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98303.0	2.6				ug/L	74699	Standard
	Be	9	480.0	19.8	**0.4216**	0.078	18.5	ug/L	8	Standard
	Al	27	74581592.0	2.0	**385.7852**	11.899	3.1	ug/L	1493	Standard
	Sc	45	57877.0	1.4				ug/L	61029	Standard
	Ti	47	6572.4	2.9	**18.4374**	0.321	1.7	ug/L	35	Standard
	V	51	8784.7	0.2	**1.0487**	0.019	1.8	ug/L	1435	Standard
	Cr	52	16164.0	1.7	**1.3092**	0.018	1.4	ug/L	7635	Standard
	Cr	53	2105.1	0.9	**2.1070**	0.052	2.5	ug/L	363	Standard
	Mn	55	1145494.2	1.1	**175.8506**	2.171	1.2	ug/L	747	Standard
	Co	59	6198.3	3.2	**0.9668**	0.035	3.6	ug/L	234	Standard
	Ni	60	8328.3	2.3	**4.6339**	0.069	1.5	ug/L	215	Standard
	Cu	65	39558.9	1.3	**22.6316**	0.139	0.6	ug/L	104	Standard
	Zn	66	37507.0	2.0	**36.2726**	0.401	1.1	ug/L	363	Standard
>	**Ge**	72	462898.1	1.3				ug/L	495142	Standard
	As	75	729.1	4.9	**0.7559**	0.035	4.6	ug/L	-50	Standard
	Se	82	97.9	9.3	**0.8168**	0.100	12.2	ug/L	14	Standard
	Se-1	77	92.7	14.7	**0.7365**	0.213	29.0	ug/L	42	Standard
>	**Ga**	71	1238.4	2.6				mg/L	42	Standard
	Rb	85	36981.8	1.6				ug/L	17	Standard
	Y	89	423750.9	0.9				ug/L	417098	Standard
>	**Rh**	103	1170.0	7.1				ug/L	13	Standard
	Mo	98	1333.1	1.3	**0.5111**	0.005	0.9	ug/L	11	Standard
	Ag	107	159.7	2.8	**0.0124**	0.001	8.3	ug/L	68	Standard
	Cd	111	50.7	12.0	**0.0171**	0.003	15.0	mg/L	2	Standard
	Cd	114	154.3	14.1	**0.0289**	0.004	13.0	ug/L	12	Standard
>	**In**	115	568573.6	2.1				ug/L	587441	Standard
	Sn	118	2385.2	8.8	**0.3269**	0.032	9.8	ug/L	497	Standard
	Sb	123	701.5	2.0	**0.1511**	0.004	2.9	ug/L	78	Standard
	Ba	135	49620.1	0.4	**25.0380**	0.447	1.8	ug/L	24	Standard
	Ce	140	69524.2	1.7				ug/L	25	Standard
>	**Tb**	159	941672.2	2.5				ug/L	939701	Standard
	Ho	165	1511.7	2.5				ug/L	10	Standard
	Tl	203	215.3	8.2	**0.0254**	0.002	6.7	ug/L	4	Standard
	Tl	205	171.7	4.4	**0.0251**	0.001	2.6	ug/L	2	Standard
	Pb	206	5244.6	2.1	**0.9485**	0.002	0.2	ug/L	240	Standard
	Pb	207	4168.9	1.9	**0.8423**	0.008	0.9	ug/L	198	Standard
	Pb	208	17565.0	0.7	**0.8990**	0.013	1.4	ug/L	883	Standard
	U	238	17410.3	1.3	**2.6228**	0.019	0.7	ug/L	3	Standard
>	**Bi**	209	504494.7	1.9				ug/L	523984	Standard

Sample ID: L1606098629
Report Date/Time: Wednesday, June 29, 2016 14:12:53
Page 1

Approved: July 04, 2016

Na	23	13.3	78.1	**161.6266**	126.012	78.0	mg/L	0	Standard
Mg	24	27959.1	0.6	**37.9939**	0.282	0.7	mg/L	37	Standard
K	39	733.4	5.5	**1.5946**	0.079	5.0	mg/L	38	Standard
Ca	43	1693.4	2.8	**110.2950**	4.030	3.7	mg/L	10	Standard
Fe	54	1507.3	3.7	**1.2090**	0.035	2.9	mg/L	179	Standard
Fe	57	1286.7	5.0	**3.5362**	0.177	5.0	mg/L	335	Standard
> **Sc-1**	45	57877.0	1.4				mg/L	61029	Standard
Cl	35	39812.2	0.7				ug/L	20195	Standard
Kr	83	2.3	24.7				ug/L	4	Standard
Br	81	15855.0	0.8				ug/L	1547	Standard
P	31	19874.7	1.1				ug/L	15286	Standard
S	34	1861.8	8.3				ug/L	1572	Standard
Sr	88	5489.3	0.8				ug/L	73	Standard
C	12	143.3	29.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	2426.6	8.1				mg/L	6	Standard
Ho-1	165	1511.7	2.5				mg/L	10	Standard
Er	166	1466.7	15.1				mg/L	13	Standard
I	127	114010.2	6.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		131.599	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.488	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098629
Report Date/Time: Wednesday, June 29, 2016 14:12:53
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.788
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.280
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098629
Report Date/Time: Wednesday, June 29, 2016 14:12:53
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098639
Sample Date/Time: Wednesday, June 29, 2016 14:13:47
Number of Replicates: 3
Autosampler Position: 330
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95709.8	2.9				ug/L	74699	Standard
	Be	9	26.7	92.5	**0.0314**	0.022	70.9	ug/L	8	Standard
	Al	27	47777.1	19.3	**0.2396**	0.041	17.2	ug/L	1493	Standard
	Sc	45	57199.5	2.6				ug/L	61029	Standard
	Ti	47	140.3	60.7	**0.2633**	0.227	86.2	ug/L	35	Standard
	V	51	2113.6	5.5	**0.0856**	0.015	17.8	ug/L	1435	Standard
	Cr	52	11519.7	1.7	**0.5382**	0.024	4.4	ug/L	7635	Standard
	Cr	53	731.7	8.2	**0.4658**	0.068	14.5	ug/L	363	Standard
	Mn	55	10046.3	0.4	**1.3019**	0.003	0.2	ug/L	747	Standard
	Co	59	277.3	4.9	**0.0049**	0.002	38.1	ug/L	234	Standard
	Ni	60	403.7	2.5	**0.0928**	0.006	6.6	ug/L	215	Standard
	Cu	65	2594.2	1.0	**1.2969**	0.011	0.8	ug/L	104	Standard
	Zn	66	5488.0	1.5	**4.5940**	0.065	1.4	ug/L	363	Standard
>	**Ge**	72	492746.6	0.5				ug/L	495142	Standard
	As	75	-23.9	159.7	**0.0381**	0.034	90.0	ug/L	-50	Standard
	Se	82	16.9	42.3	**0.0162**	0.065	398.3	ug/L	14	Standard
	Se-1	77	44.7	15.2	**-0.0080**	0.091	1136.4	ug/L	42	Standard
>	**Ga**	71	35.0	24.7				mg/L	42	Standard
	Rb	85	228.3	6.7				ug/L	17	Standard
	Y	89	426313.2	2.4				ug/L	417098	Standard
>	**Rh**	103	20.0	66.1				ug/L	13	Standard
	Mo	98	146.2	8.6	**0.0487**	0.005	10.8	ug/L	11	Standard
	Ag	107	84.7	10.6	**-0.0000**	0.001	3427.7	ug/L	68	Standard
	Cd	111	14.4	26.2	**-0.0011**	0.002	163.5	mg/L	2	Standard
	Cd	114	36.0	7.5	**0.0048**	0.001	11.2	ug/L	12	Standard
>	**In**	115	592771.3	1.1				ug/L	587441	Standard
	Sn	118	1206.7	2.0	**0.1101**	0.007	6.0	ug/L	497	Standard
	Sb	123	2747.6	2.7	**0.5790**	0.016	2.8	ug/L	78	Standard
	Ba	135	327.7	1.8	**0.1426**	0.002	1.3	ug/L	24	Standard
	Ce	140	215.0	46.7				ug/L	25	Standard
>	**Tb**	159	942434.8	1.8				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	30.7	28.5	**0.0019**	0.001	54.2	ug/L	4	Standard
	Tl	205	31.7	59.8	**0.0045**	0.003	57.0	ug/L	2	Standard
	Pb	206	347.7	8.5	**-0.0077**	0.005	63.4	ug/L	240	Standard
	Pb	207	296.0	3.8	**-0.0091**	0.002	22.4	ug/L	198	Standard
	Pb	208	1207.0	5.0	**-0.0113**	0.003	23.3	ug/L	883	Standard
	U	238	16.7	21.1	**-0.0021**	0.000	23.4	ug/L	3	Standard
>	**Bi**	209	536687.4	0.8				ug/L	523984	Standard

Sample ID: L1606098639
Report Date/Time: Wednesday, June 29, 2016 14:16:04
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	78.3	9.8	**0.0427**	0.008	19.8	mg/L	37	Standard
K	39	33.3	56.8	**-0.0347**	0.047	136.1	mg/L	38	Standard
Ca	43	38.3	27.2	**0.0551**	0.672	1218.0	mg/L	10	Standard
Fe	54	152.5	13.6	**-0.0001**	0.021	24429.9	mg/L	179	Standard
Fe	57	270.0	10.3	**-0.0961**	0.083	86.0	mg/L	335	Standard
Sc-1	45	57199.5	2.6				mg/L	61029	Standard
Cl	35	30633.6	0.6				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	2093.5	12.2				ug/L	1547	Standard
P	31	17446.7	2.4				ug/L	15286	Standard
S	34	1971.8	2.9				ug/L	1572	Standard
Sr	88	48.3	26.0				ug/L	73	Standard
C	12	116.7	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	9.0	108.0				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	4304.0	50.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		128.128	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.516	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098639
Report Date/Time: Wednesday, June 29, 2016 14:16:04
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.907
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.424
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098641
Sample Date/Time: Wednesday, June 29, 2016 14:16:59
Number of Replicates: 3
Autosampler Position: 331
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87267.4	3.2				ug/L	74699	Standard
	Be	9	11.7	49.5	**0.0189**	0.005	28.8	ug/L	8	Standard
	Al	27	605511.0	1.7	**3.5149**	0.083	2.4	ug/L	1493	Standard
	Sc	45	54390.8	2.9				ug/L	61029	Standard
	Ti	47	93.0	12.4	**0.1442**	0.032	22.4	ug/L	35	Standard
	V	51	1794.1	5.9	**0.0496**	0.015	30.2	ug/L	1435	Standard
	Cr	52	10365.2	1.5	**0.4195**	0.024	5.7	ug/L	7635	Standard
	Cr	53	613.3	8.2	**0.3551**	0.058	16.5	ug/L	363	Standard
	Mn	55	4513.0	1.6	**0.5208**	0.012	2.3	ug/L	747	Standard
	Co	59	310.3	5.3	**0.0112**	0.002	21.0	ug/L	234	Standard
	Ni	60	490.0	6.7	**0.1461**	0.017	11.6	ug/L	215	Standard
	Cu	65	1018.0	2.6	**0.4608**	0.017	3.7	ug/L	104	Standard
	Zn	66	3968.5	2.5	**3.2944**	0.091	2.8	ug/L	363	Standard
>	**Ge**	72	479837.7	0.4				ug/L	495142	Standard
	As	75	-18.8	39.3	**0.0423**	0.007	15.9	ug/L	-50	Standard
	Se	82	16.1	20.6	**0.0139**	0.032	229.1	ug/L	14	Standard
	Se-1	77	51.0	14.8	**0.0980**	0.104	105.7	ug/L	42	Standard
>	**Ga**	71	46.7	22.3				mg/L	42	Standard
	Rb	85	2146.8	6.0				ug/L	17	Standard
	Y	89	405082.7	1.5				ug/L	417098	Standard
>	**Rh**	103	10.0	86.6				ug/L	13	Standard
	Mo	98	183.9	8.1	**0.0672**	0.006	9.5	ug/L	11	Standard
	Ag	107	85.0	8.5	**0.0009**	0.001	123.2	ug/L	68	Standard
	Cd	111	58.3	9.8	**0.0215**	0.003	12.3	mg/L	2	Standard
	Cd	114	162.0	10.0	**0.0312**	0.003	9.6	ug/L	12	Standard
>	**In**	115	556306.9	1.6				ug/L	587441	Standard
	Sn	118	983.4	6.5	**0.0831**	0.013	16.0	ug/L	497	Standard
	Sb	123	290.8	14.6	**0.0617**	0.011	17.2	ug/L	78	Standard
	Ba	135	1301.4	2.0	**0.6554**	0.007	1.1	ug/L	24	Standard
	Ce	140	1251.7	2.7				ug/L	25	Standard
>	**Tb**	159	894371.7	2.0				ug/L	939701	Standard
	Ho	165	23.3	61.9				ug/L	10	Standard
	Tl	203	26.7	21.3	**0.0015**	0.001	46.3	ug/L	4	Standard
	Tl	205	18.3	68.6	**0.0028**	0.002	63.1	ug/L	2	Standard
	Pb	206	368.7	5.0	**-0.0008**	0.004	496.5	ug/L	240	Standard
	Pb	207	301.3	5.5	**-0.0052**	0.004	82.3	ug/L	198	Standard
	Pb	208	1198.4	1.5	**-0.0089**	0.002	20.6	ug/L	883	Standard
	U	238	7.7	32.8	**-0.0033**	0.000	11.3	ug/L	3	Standard
>	**Bi**	209	513924.4	1.5				ug/L	523984	Standard

Sample ID: L1606098641
Report Date/Time: Wednesday, June 29, 2016 14:19:15
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	638.3	13.6	**0.8624**	0.151	17.5	mg/L	37	Standard
K	39	86.7	3.3	**0.1010**	0.012	12.3	mg/L	38	Standard
Ca	43	31.7	48.2	**-0.2871**	1.074	374.0	mg/L	10	Standard
Fe	54	129.1	21.5	**-0.0150**	0.029	192.4	mg/L	179	Standard
Fe	57	250.0	12.2	**-0.1181**	0.141	119.7	mg/L	335	Standard
Sc-1	45	54390.8	2.9				mg/L	61029	Standard
Cl	35	28123.4	3.1				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1706.8	4.2				ug/L	1547	Standard
P	31	17456.7	1.2				ug/L	15286	Standard
S	34	1960.1	4.3				ug/L	1572	Standard
Sr	88	86.7	3.3				ug/L	73	Standard
C	12	86.7	35.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	22.9	67.6				mg/L	6	Standard
Ho-1	165	23.3	61.9				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	6631.5	5.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		116.826	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.909	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098641
Report Date/Time: Wednesday, June 29, 2016 14:19:15
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.700
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.080
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102602
Sample Date/Time: Wednesday, June 29, 2016 14:51:20
Number of Replicates: 3
Autosampler Position: 332
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98959.1	4.1				ug/L	74699	Standard
	Be	9	893.4	7.1	**0.7730**	0.024	3.1	ug/L	8	Standard
	Al	27	86838575.1	3.0	**446.1622**	5.675	1.3	ug/L	1493	Standard
	Sc	45	59841.3	1.2				ug/L	61029	Standard
	Ti	47	724.7	2.7	**1.9675**	0.086	4.4	ug/L	35	Standard
	V	51	-99401.9	27.0	**-14.4927**	3.863	26.7	ug/L	1435	Standard
	Cr	52	14286817.6	0.7	**2092.1702**	38.595	1.8	ug/L	7635	Standard
	Cr	53	1819264.9	1.0	**2136.8631**	30.527	1.4	ug/L	363	Standard
	Mn	55	9369393.8	2.3	**1460.7209**	18.166	1.2	ug/L	747	Standard
	Co	59	953653.0	0.8	**156.7618**	2.325	1.5	ug/L	234	Standard
	Ni	60	10094931.6	0.5	**5854.0954**	89.919	1.5	ug/L	215	Standard
	Cu	65	555284.9	0.4	**323.8508**	6.370	2.0	ug/L	104	Standard
	Zn	66	64094.6	1.0	**63.2545**	1.058	1.7	ug/L	363	Standard
>	**Ge**	72	456115.5	1.9				ug/L	495142	Standard
	As	75	3627.2	1.9	**3.5767**	0.135	3.8	ug/L	-50	Standard
	Se	82	1754.2	3.0	**17.2111**	0.289	1.7	ug/L	14	Standard
	Se-1	77	3450.1	4.6	**50.7001**	3.400	6.7	ug/L	42	Standard
>	**Ga**	71	493.3	9.8				mg/L	42	Standard
	Rb	85	15579.7	1.3				ug/L	17	Standard
	Y	89	510259.3	1.5				ug/L	417098	Standard
>	**Rh**	103	536.7	10.8				ug/L	13	Standard
	Mo	98	52406.3	2.2	**22.0721**	0.200	0.9	ug/L	11	Standard
	Ag	107	797.7	2.2	**0.1246**	0.004	3.1	ug/L	68	Standard
	Cd	111	4820.0	1.8	**2.5808**	0.025	1.0	mg/L	2	Standard
	Cd	114	11654.3	2.8	**2.5494**	0.039	1.5	ug/L	12	Standard
>	**In**	115	523117.8	1.3				ug/L	587441	Standard
	Sn	118	2368.5	5.4	**0.3606**	0.023	6.4	ug/L	497	Standard
	Sb	123	2240.5	1.3	**0.5347**	0.001	0.2	ug/L	78	Standard
	Ba	135	135783.1	0.9	**74.4848**	0.361	0.5	ug/L	24	Standard
	Ce	140	104687.1	1.1				ug/L	25	Standard
>	**Tb**	159	898543.7	0.7				ug/L	939701	Standard
	Ho	165	6684.8	2.2				ug/L	10	Standard
	Tl	203	178.0	4.1	**0.0225**	0.001	5.6	ug/L	4	Standard
	Tl	205	203.3	37.2	**0.0321**	0.012	38.4	ug/L	2	Standard
	Pb	206	1304.4	5.0	**0.2026**	0.016	8.1	ug/L	240	Standard
	Pb	207	1055.7	2.5	**0.1793**	0.004	2.5	ug/L	198	Standard
	Pb	208	4341.2	2.0	**0.1854**	0.004	2.0	ug/L	883	Standard
	U	238	4585.0	0.8	**0.7423**	0.007	0.9	ug/L	3	Standard
>	**Bi**	209	467454.5	1.3				ug/L	523984	Standard

Sample ID: L1606102602
Report Date/Time: Wednesday, June 29, 2016 14:53:37
Page 1

Approved: July 04, 2016

Na	23	5.0	0.0	**58.7812**	0.729	1.2	mg/L	0	Standard
Mg	24	188890.8	2.9	**248.6881**	10.353	4.2	mg/L	37	Standard
K	39	2185.2	4.4	**4.8106**	0.175	3.6	mg/L	38	Standard
Ca	43	1135.0	6.5	**70.5642**	3.873	5.5	mg/L	10	Standard
Fe	54	9878.2	4.8	**8.4031**	0.399	4.8	mg/L	179	Standard
Fe	57	3267.0	5.5	**10.2558**	0.748	7.3	mg/L	335	Standard
> **Sc-1**	45	59841.3	1.2				mg/L	61029	Standard
Cl	35	28240.3	1.1				ug/L	20195	Standard
Kr	83	2.7	142.0				ug/L	4	Standard
Br	81	339155.5	2.6				ug/L	1547	Standard
P	31	29607.2	0.7				ug/L	15286	Standard
S	34	2378.5	1.9				ug/L	1572	Standard
Sr	88	1895.1	7.4				ug/L	73	Standard
C	12	120.0	28.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	86.7	35.3				mg/L	3	Standard
Dy	164	9299.8	2.7				mg/L	6	Standard
Ho-1	165	6684.8	2.2				mg/L	10	Standard
Er	166	6331.3	1.7				mg/L	13	Standard
I	127	153619.9	12.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.478	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.118	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606102602
Report Date/Time: Wednesday, June 29, 2016 14:53:37
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	89.050
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	89.212
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
V 51 Lower	V	51	

Sample ID: L1606102602
Report Date/Time: Wednesday, June 29, 2016 14:53:37
Page 3

Approved: July 04, 2016

Cr 52 Upper, S, EEE	Cr	52
Cr 53 Upper, S, EEE	Cr	53
Mn 55 Upper, S, EEE	Mn	55
Co 59 Upper, S, EEE	Co	59
Ni 60 Upper, S, EEE	Ni	60
Cu 65 Upper, S, EEE	Cu	65

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102604
Sample Date/Time: Wednesday, June 29, 2016 14:54:31
Number of Replicates: 3
Autosampler Position: 333
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82386.8	1.6				ug/L	74699	Standard
	Be	9	41.7	50.0	**0.0505**	0.021	41.9	ug/L	8	Standard
	Al	27	6418180.7	1.0	**39.5946**	1.017	2.6	ug/L	1493	Standard
	Sc	45	68191.6	3.4				ug/L	61029	Standard
	Ti	47	2431.5	5.8	**6.2763**	0.263	4.2	ug/L	35	Standard
	V	51	8932.1	2.8	**0.9822**	0.034	3.4	ug/L	1435	Standard
	Cr	52	19971.2	5.8	**1.6585**	0.113	6.8	ug/L	7635	Standard
	Cr	53	16018.5	1.2	**16.9336**	0.479	2.8	ug/L	363	Standard
	Mn	55	134583.1	0.5	**19.1035**	0.326	1.7	ug/L	747	Standard
	Co	59	5374.0	2.6	**0.7730**	0.008	1.1	ug/L	234	Standard
	Ni	60	7430.2	8.6	**3.8249**	0.274	7.2	ug/L	215	Standard
	Cu	65	1488.1	5.9	**0.6921**	0.035	5.0	ug/L	104	Standard
	Zn	66	5451.0	0.8	**4.5156**	0.102	2.3	ug/L	363	Standard
>	**Ge**	72	497254.0	1.7				ug/L	495142	Standard
	As	75	427.8	16.4	**0.4399**	0.063	14.3	ug/L	-50	Standard
	Se	82	161.6	5.9	**1.3278**	0.065	4.9	ug/L	14	Standard
	Se-1	77	695.3	4.1	**8.8552**	0.225	2.5	ug/L	42	Standard
>	**Ga**	71	420.0	2.4				mg/L	42	Standard
	Rb	85	8292.3	2.6				ug/L	17	Standard
	Y	89	430205.9	2.4				ug/L	417098	Standard
>	**Rh**	103	40.0	12.5				ug/L	13	Standard
	Mo	98	177.9	8.2	**0.0624**	0.005	8.3	ug/L	11	Standard
	Ag	107	97.3	5.7	**0.0023**	0.001	38.0	ug/L	68	Standard
	Cd	111	330.0	5.3	**0.1533**	0.008	5.4	mg/L	2	Standard
	Cd	114	777.4	4.0	**0.1527**	0.007	4.3	ug/L	12	Standard
>	**In**	115	575297.3	2.0				ug/L	587441	Standard
	Sn	118	1836.8	3.5	**0.2263**	0.005	2.3	ug/L	497	Standard
	Sb	123	541.3	2.5	**0.1143**	0.005	4.2	ug/L	78	Standard
	Ba	135	259403.2	1.0	**129.4148**	1.250	1.0	ug/L	24	Standard
	Ce	140	48153.1	1.6				ug/L	25	Standard
>	**Tb**	159	953820.3	0.6				ug/L	939701	Standard
	Ho	165	763.4	3.9				ug/L	10	Standard
	Tl	203	147.7	11.5	**0.0162**	0.002	12.5	ug/L	4	Standard
	Tl	205	135.0	25.9	**0.0192**	0.005	26.2	ug/L	2	Standard
	Pb	206	2333.8	1.6	**0.3679**	0.004	1.0	ug/L	240	Standard
	Pb	207	1897.5	1.5	**0.3318**	0.006	1.8	ug/L	198	Standard
	Pb	208	7911.8	1.2	**0.3500**	0.006	1.7	ug/L	883	Standard
	U	238	272.0	3.5	**0.0353**	0.002	4.3	ug/L	3	Standard
>	**Bi**	209	521315.2	0.9				ug/L	523984	Standard

Sample ID: L1606102604
Report Date/Time: Wednesday, June 29, 2016 14:56:48
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	45454.8	3.6	**52.4459**	0.539	1.0	mg/L	37	Standard
K	39	88.3	3.3	**0.0609**	0.011	18.4	mg/L	38	Standard
Ca	43	100.0	8.7	**3.1440**	0.659	21.0	mg/L	10	Standard
Fe	54	556.2	7.6	**0.2841**	0.037	12.9	mg/L	179	Standard
Fe	57	410.0	16.9	**0.1734**	0.212	122.1	mg/L	335	Standard
> **Sc-1**	45	68191.6	3.4				mg/L	61029	Standard
Cl	35	26725.6	1.1				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	35244.3	3.6				ug/L	1547	Standard
P	31	35015.4	1.7				ug/L	15286	Standard
S	34	2091.8	5.9				ug/L	1572	Standard
Sr	88	225.0	20.0				ug/L	73	Standard
C	12	103.3	39.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	93.3	12.4				mg/L	3	Standard
Dy	164	1034.1	12.7				mg/L	6	Standard
Ho-1	165	763.4	3.9				mg/L	10	Standard
Er	166	680.0	15.4				mg/L	13	Standard
I	127	80476.5	1.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.292	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.426	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606102604
Report Date/Time: Wednesday, June 29, 2016 14:56:48
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.933
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.491
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Ba 135 Upper, S, EEE	Ba	135	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102606
Sample Date/Time: Wednesday, June 29, 2016 14:57:42
Number of Replicates: 3
Autosampler Position: 334
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93356.7	0.9				ug/L	74699	Standard
	Be	9	136.7	43.6	**0.1318**	0.054	40.9	ug/L	8	Standard
	Al	27	4727947.0	2.8	**25.7324**	0.870	3.4	ug/L	1493	Standard
	Sc	45	71265.6	1.4				ug/L	61029	Standard
	Ti	47	7167.7	1.1	**18.5395**	0.116	0.6	ug/L	35	Standard
	V	51	33631.6	0.4	**4.1963**	0.093	2.2	ug/L	1435	Standard
	Cr	52	43684.3	3.9	**4.7878**	0.170	3.6	ug/L	7635	Standard
	Cr	53	18915.1	0.4	**19.8540**	0.330	1.7	ug/L	363	Standard
	Mn	55	527300.2	1.1	**74.5358**	0.480	0.6	ug/L	747	Standard
	Co	59	47125.2	1.5	**6.9996**	0.016	0.2	ug/L	234	Standard
	Ni	60	57578.9	1.0	**30.2025**	0.274	0.9	ug/L	215	Standard
	Cu	65	3787.5	1.8	**1.9028**	0.007	0.4	ug/L	104	Standard
	Zn	66	29686.7	1.5	**26.3448**	0.262	1.0	ug/L	363	Standard
>	**Ge**	72	502153.4	1.7				ug/L	495142	Standard
	As	75	1552.5	5.8	**1.4257**	0.058	4.0	ug/L	-50	Standard
	Se	82	328.6	4.9	**2.8136**	0.101	3.6	ug/L	14	Standard
	Se-1	77	866.0	3.6	**11.0672**	0.235	2.1	ug/L	42	Standard
>	**Ga**	71	2006.8	3.5				mg/L	42	Standard
	Rb	85	28431.7	2.8				ug/L	17	Standard
	Y	89	467809.8	1.5				ug/L	417098	Standard
>	**Rh**	103	28.3	36.7				ug/L	13	Standard
	Mo	98	2788.7	2.4	**1.0189**	0.018	1.8	ug/L	11	Standard
	Ag	107	187.0	8.5	**0.0152**	0.002	14.7	ug/L	68	Standard
	Cd	111	184.5	8.6	**0.0785**	0.007	8.5	mg/L	2	Standard
	Cd	114	502.9	12.6	**0.0940**	0.013	13.9	ug/L	12	Standard
>	**In**	115	599816.0	1.2				ug/L	587441	Standard
	Sn	118	1385.1	2.5	**0.1375**	0.003	2.2	ug/L	497	Standard
	Sb	123	315.5	7.5	**0.0620**	0.005	7.5	ug/L	78	Standard
	Ba	135	37805.8	2.1	**18.0728**	0.162	0.9	ug/L	24	Standard
	Ce	140	83749.2	1.1				ug/L	25	Standard
>	**Tb**	159	971524.4	1.2				ug/L	939701	Standard
	Ho	165	1393.4	2.1				ug/L	10	Standard
	Tl	203	289.7	7.1	**0.0339**	0.002	6.5	ug/L	4	Standard
	Tl	205	258.3	11.0	**0.0368**	0.004	11.7	ug/L	2	Standard
	Pb	206	3853.2	1.2	**0.6604**	0.004	0.6	ug/L	240	Standard
	Pb	207	3010.0	1.5	**0.5732**	0.009	1.6	ug/L	198	Standard
	Pb	208	13337.3	2.3	**0.6473**	0.013	1.9	ug/L	883	Standard
	U	238	859.0	4.4	**0.1221**	0.005	4.0	ug/L	3	Standard
>	**Bi**	209	516635.2	0.9				ug/L	523984	Standard

Sample ID: L1606102606
Report Date/Time: Wednesday, June 29, 2016 14:59:59
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.3563**	28.321	173.2	mg/L	0	Standard
Mg	24	103702.1	2.0	**114.5988**	3.552	3.1	mg/L	37	Standard
K	39	251.7	12.0	**0.3627**	0.062	17.2	mg/L	38	Standard
Ca	43	61.7	26.1	**0.8208**	0.922	112.4	mg/L	10	Standard
Fe	54	1412.4	2.3	**0.8871**	0.015	1.7	mg/L	179	Standard
Fe	57	645.0	2.8	**0.8030**	0.026	3.3	mg/L	335	Standard
Sc-1	45	71265.6	1.4				mg/L	61029	Standard
Cl	35	28407.3	0.4				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	49253.3	3.6				ug/L	1547	Standard
P	31	41409.9	2.5				ug/L	15286	Standard
S	34	2068.5	4.4				ug/L	1572	Standard
Sr	88	155.0	22.6				ug/L	73	Standard
C	12	133.3	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	57.3				mg/L	3	Standard
Dy	164	2353.9	2.8				mg/L	6	Standard
Ho-1	165	1393.4	2.1				mg/L	10	Standard
Er	166	1453.4	9.2				mg/L	13	Standard
I	127	3067238.7	11.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		124.978	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		101.416	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606102606
Report Date/Time: Wednesday, June 29, 2016 14:59:59
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.107
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.597
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102607
Sample Date/Time: Wednesday, June 29, 2016 15:00:54
Number of Replicates: 3
Autosampler Position: 335
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86025.2	1.8				ug/L	74699	Standard
	Be	9	83.3	29.6	**0.0901**	0.026	28.3	ug/L	8	Standard
	Al	27	3430436.9	1.2	**20.2582**	0.272	1.3	ug/L	1493	Standard
	Sc	45	67818.2	2.7				ug/L	61029	Standard
	Ti	47	2650.2	2.1	**6.1960**	0.095	1.5	ug/L	35	Standard
	V	51	9706.1	1.4	**0.9634**	0.025	2.6	ug/L	1435	Standard
	Cr	52	14237.0	2.0	**0.7086**	0.035	4.9	ug/L	7635	Standard
	Cr	53	2521.9	11.8	**2.1295**	0.275	12.9	ug/L	363	Standard
	Mn	55	50752.2	3.1	**6.4246**	0.167	2.6	ug/L	747	Standard
	Co	59	8541.7	4.1	**1.1290**	0.040	3.6	ug/L	234	Standard
	Ni	60	2905.6	5.8	**1.2753**	0.074	5.8	ug/L	215	Standard
	Cu	65	1384.7	3.1	**0.5672**	0.017	3.0	ug/L	104	Standard
	Zn	66	5145.2	1.6	**3.7928**	0.039	1.0	ug/L	363	Standard
>	**Ge**	72	549096.7	0.7				ug/L	495142	Standard
	As	75	307.3	24.2	**0.3069**	0.059	19.1	ug/L	-50	Standard
	Se	82	131.2	8.7	**0.9395**	0.088	9.3	ug/L	14	Standard
	Se-1	77	177.7	11.5	**1.5705**	0.241	15.4	ug/L	42	Standard
>	**Ga**	71	765.0	18.4				mg/L	42	Standard
	Rb	85	11574.4	1.2				ug/L	17	Standard
	Y	89	476523.2	1.2				ug/L	417098	Standard
>	**Rh**	103	40.0	54.5				ug/L	13	Standard
	Mo	98	552.0	3.2	**0.1876**	0.008	4.1	ug/L	11	Standard
	Ag	107	103.3	5.6	**0.0019**	0.001	51.8	ug/L	68	Standard
	Cd	111	131.7	9.5	**0.0509**	0.006	12.0	mg/L	2	Standard
	Cd	114	303.7	9.9	**0.0531**	0.005	10.1	ug/L	12	Standard
>	**In**	115	629575.3	1.6				ug/L	587441	Standard
	Sn	118	1851.8	1.5	**0.2011**	0.002	0.9	ug/L	497	Standard
	Sb	123	475.2	2.5	**0.0908**	0.004	4.2	ug/L	78	Standard
	Ba	135	119108.2	2.3	**54.2912**	1.481	2.7	ug/L	24	Standard
	Ce	140	65755.7	1.8				ug/L	25	Standard
>	**Tb**	159	1004571.7	2.5				ug/L	939701	Standard
	Ho	165	1471.7	8.7				ug/L	10	Standard
	Tl	203	123.3	5.2	**0.0121**	0.001	4.7	ug/L	4	Standard
	Tl	205	98.3	10.6	**0.0129**	0.002	11.8	ug/L	2	Standard
	Pb	206	2854.9	1.5	**0.4229**	0.014	3.4	ug/L	240	Standard
	Pb	207	2325.8	2.5	**0.3831**	0.018	4.6	ug/L	198	Standard
	Pb	208	9906.6	0.9	**0.4144**	0.011	2.7	ug/L	883	Standard
	U	238	660.3	1.2	**0.0843**	0.002	2.3	ug/L	3	Standard
>	**Bi**	209	566874.3	1.4				ug/L	523984	Standard

Sample ID: L1606102607
Report Date/Time: Wednesday, June 29, 2016 15:03:11
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	9059.4	5.1	**10.4668**	0.674	6.4	mg/L	37	Standard
K	39	175.0	7.6	**0.2337**	0.017	7.4	mg/L	38	Standard
Ca	43	76.7	10.0	**1.8315**	0.359	19.6	mg/L	10	Standard
Fe	54	778.8	6.8	**0.4564**	0.046	10.1	mg/L	179	Standard
Fe	57	450.0	7.8	**0.3039**	0.129	42.4	mg/L	335	Standard
> **Sc-1**	45	67818.2	2.7				mg/L	61029	Standard
Cl	35	27265.2	1.0				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	4797.4	2.1				ug/L	1547	Standard
P	31	22328.1	2.9				ug/L	15286	Standard
S	34	1868.4	11.0				ug/L	1572	Standard
Sr	88	170.0	23.0				ug/L	73	Standard
C	12	116.7	4.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	2276.8	3.2				mg/L	6	Standard
Ho-1	165	1471.7	8.7				mg/L	10	Standard
Er	166	1323.4	4.9				mg/L	13	Standard
I	127	142556.3	31.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		115.163	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.897	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606102607
Report Date/Time: Wednesday, June 29, 2016 15:03:11
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		107.172
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.185
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 15:04:07
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	83434.4	4.1				ug/L	74699	Standard
	Be	9	52151.3	1.6	**53.0799**	1.438	2.7	ug/L	8	Standard
	Al	27	8001829.5	1.2	**48.7762**	1.563	3.2	ug/L	1493	Standard
	Sc	45	63472.7	2.1				ug/L	61029	Standard
	Ti	47	41544.5	0.5	**100.8598**	1.507	1.5	ug/L	35	Standard
	V	51	430823.0	0.6	**52.3478**	0.567	1.1	ug/L	1435	Standard
	Cr	52	414183.8	1.0	**50.4520**	0.020	0.0	ug/L	7635	Standard
	Cr	53	51435.5	2.1	**50.9262**	0.597	1.2	ug/L	363	Standard
	Mn	55	402509.0	2.2	**53.1108**	1.647	3.1	ug/L	747	Standard
	Co	59	395807.1	0.7	**55.1753**	0.902	1.6	ug/L	234	Standard
	Ni	60	102515.7	1.1	**50.3079**	0.215	0.4	ug/L	215	Standard
	Cu	65	102297.5	1.1	**50.5232**	0.685	1.4	ug/L	104	Standard
	Zn	66	61288.7	1.4	**51.2272**	0.860	1.7	ug/L	363	Standard
>	Ge	72	537577.7	1.0				ug/L	495142	Standard
	As	75	61234.0	0.5	**50.4168**	0.738	1.5	ug/L	-50	Standard
	Se	82	6132.0	1.1	**51.3300**	1.089	2.1	ug/L	14	Standard
	Se-1	77	4299.6	0.9	**53.6055**	0.977	1.8	ug/L	42	Standard
>	Ga	71	78.3	42.5				mg/L	42	Standard
	Rb	85	866.7	7.4				ug/L	17	Standard
	Y	89	460058.1	1.7				ug/L	417098	Standard
>	Rh	103	33.3	31.2				ug/L	13	Standard
	Mo	98	295506.5	0.5	**103.9241**	2.114	2.0	ug/L	11	Standard
	Ag	107	348188.0	0.5	**50.0657**	0.748	1.5	ug/L	68	Standard
	Cd	111	111062.0	1.3	**49.7777**	0.346	0.7	mg/L	2	Standard
	Cd	114	271614.5	0.9	**49.6432**	0.990	2.0	ug/L	12	Standard
>	In	115	626816.0	1.9				ug/L	587441	Standard
	Sn	118	313548.1	1.3	**50.2246**	1.384	2.8	ug/L	497	Standard
	Sb	123	246409.0	0.1	**49.4692**	0.867	1.8	ug/L	78	Standard
	Ba	135	105775.6	0.8	**48.4328**	1.255	2.6	ug/L	24	Standard
	Ce	140	66.7	48.8				ug/L	25	Standard
>	Tb	159	996778.4	1.8				ug/L	939701	Standard
	Ho	165	16.7	34.6				ug/L	10	Standard
	Tl	203	442598.6	0.7	**50.9461**	0.764	1.5	ug/L	4	Standard
	Tl	205	402531.2	1.7	**53.3816**	0.705	1.3	ug/L	2	Standard
	Pb	206	287928.0	0.2	**51.0713**	0.393	0.8	ug/L	240	Standard
	Pb	207	257921.1	0.2	**51.4941**	0.566	1.1	ug/L	198	Standard
	Pb	208	1040092.9	0.4	**52.5571**	0.545	1.0	ug/L	883	Standard
	U	238	381264.5	1.2	**52.5495**	0.479	0.9	ug/L	3	Standard
>	Bi	209	552262.9	0.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:06:24
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4305.6	5.5	**5.2771**	0.213	4.0	mg/L	37	Standard
K	39	2275.2	2.4	**4.7239**	0.217	4.6	mg/L	38	Standard
Ca	43	126.7	42.6	**5.1579**	3.228	62.6	mg/L	10	Standard
Fe	54	6415.2	4.5	**5.0956**	0.328	6.4	mg/L	179	Standard
Fe	57	1851.8	4.1	**4.9827**	0.373	7.5	mg/L	335	Standard
> **Sc-1**	45	63472.7	2.1				mg/L	61029	Standard
Cl	35	28119.4	1.3				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2310.2	13.5				ug/L	1547	Standard
P	31	19991.5	1.1				ug/L	15286	Standard
S	34	2273.5	5.1				ug/L	1572	Standard
Sr	88	71.7	10.7				ug/L	73	Standard
C	12	66.7	8.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	8.6	116.8				mg/L	6	Standard
Ho-1	165	16.7	34.6				mg/L	10	Standard
Er	166	30.0	0.0				mg/L	13	Standard
I	127	25116.6	23.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	106.160		
Al	27	97.552		
Sc	45			
Ti	47	100.860		
V	51	104.696		
Cr	52	100.904		
Cr	53			
Mn	55	106.222		
Co	59	110.351		
Ni	60	100.616		
Cu	65	101.046		
Zn	66	102.454		
> Ge	72		108.570	
As	75	100.834		
Se	82	102.660		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:06:24
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	103.924		
Ag	107	100.131		
Cd	111	99.555		
Cd	114			
In	115			106.703
Sn	118	100.449		
Sb	123	98.938		
Ba	135	96.866		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.892		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.114		
U	238	105.099		
Bi	209			105.397
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:06:24
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 15:07:18
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85012.9	2.0				ug/L	74699	Standard
	Be	9	28.3	27.0	**0.0359**	0.008	21.1	ug/L	8	Standard
	Al	27	3885.6	39.7	**0.0100**	0.010	97.9	ug/L	1493	Standard
	Sc	45	66216.1	1.0				ug/L	61029	Standard
	Ti	47	42.3	11.2	**-0.0066**	0.011	167.9	ug/L	35	Standard
	V	51	1610.8	4.5	**0.0001**	0.010	18188.0	ug/L	1435	Standard
	Cr	52	7733.0	4.2	**-0.0667**	0.042	62.7	ug/L	7635	Standard
	Cr	53	656.7	7.6	**0.3218**	0.049	15.4	ug/L	363	Standard
	Mn	55	934.4	2.8	**-0.0250**	0.004	15.9	ug/L	747	Standard
	Co	59	270.3	16.3	**0.0003**	0.006	2278.1	ug/L	234	Standard
	Ni	60	353.0	46.0	**0.0492**	0.080	162.9	ug/L	215	Standard
	Cu	65	160.7	5.6	**-0.0245**	0.004	18.2	ug/L	104	Standard
	Zn	66	427.3	1.4	**-0.0957**	0.006	6.2	ug/L	363	Standard
>	**Ge**	72	539852.0	0.5				ug/L	495142	Standard
	As	75	-35.2	72.3	**0.0308**	0.021	67.1	ug/L	-50	Standard
	Se	82	23.4	13.4	**0.0577**	0.027	47.1	ug/L	14	Standard
	Se-1	77	65.0	15.6	**0.1944**	0.131	67.4	ug/L	42	Standard
>	**Ga**	71	48.3	15.8				mg/L	42	Standard
	Rb	85	13.3	21.7				ug/L	17	Standard
	Y	89	460334.1	1.4				ug/L	417098	Standard
>	**Rh**	103	15.0	57.7				ug/L	13	Standard
	Mo	98	325.6	30.7	**0.1063**	0.035	33.0	ug/L	11	Standard
	Ag	107	121.0	14.9	**0.0041**	0.003	62.9	ug/L	68	Standard
	Cd	111	13.2	31.2	**-0.0021**	0.002	85.0	mg/L	2	Standard
	Cd	114	51.9	43.2	**0.0072**	0.004	56.5	ug/L	12	Standard
>	**In**	115	641477.2	0.6				ug/L	587441	Standard
	Sn	118	860.0	23.5	**0.0400**	0.031	77.0	ug/L	497	Standard
	Sb	123	396.3	20.9	**0.0735**	0.016	21.4	ug/L	78	Standard
	Ba	135	45.3	20.4	**0.0043**	0.004	99.1	ug/L	24	Standard
	Ce	140	36.7	15.7				ug/L	25	Standard
>	**Tb**	159	1015779.3	0.6				ug/L	939701	Standard
	Ho	165	5.0	100.0				ug/L	10	Standard
	Tl	203	76.0	35.5	**0.0066**	0.003	43.8	ug/L	4	Standard
	Tl	205	63.3	18.2	**0.0082**	0.001	17.5	ug/L	2	Standard
	Pb	206	309.7	4.7	**-0.0185**	0.002	12.2	ug/L	240	Standard
	Pb	207	263.0	6.6	**-0.0197**	0.003	14.7	ug/L	198	Standard
	Pb	208	1081.0	8.8	**-0.0217**	0.004	19.1	ug/L	883	Standard
	U	238	55.3	39.0	**0.0029**	0.003	97.3	ug/L	3	Standard
>	**Bi**	209	576716.4	0.8				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:09:35
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.8612**	30.928	173.2	mg/L	0	Standard
Mg	24	48.3	39.2	**-0.0075**	0.023	309.1	mg/L	37	Standard
K	39	65.0	38.5	**0.0181**	0.050	275.9	mg/L	38	Standard
Ca	43	31.7	9.1	**-0.6795**	0.172	25.4	mg/L	10	Standard
Fe	54	165.7	15.0	**-0.0089**	0.019	210.6	mg/L	179	Standard
Fe	57	258.3	12.9	**-0.2654**	0.099	37.5	mg/L	335	Standard
Sc-1	45	66216.1	1.0				mg/L	61029	Standard
Cl	35	27856.9	0.3				ug/L	20195	Standard
Kr	83	4.0	43.3				ug/L	4	Standard
Br	81	1983.5	3.1				ug/L	1547	Standard
P	31	20308.6	3.1				ug/L	15286	Standard
S	34	2271.8	3.7				ug/L	1572	Standard
Sr	88	51.7	22.3				ug/L	73	Standard
C	12	73.3	34.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	20.0	86.6				mg/L	6	Standard
Ho-1	165	5.0	100.0				mg/L	10	Standard
Er	166	0.0					mg/L	13	Standard
I	127	13137.4	11.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		109.030	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:09:35
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.199
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.064
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:09:35
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBS M5 WG573099-02
Sample Date/Time: Wednesday, June 29, 2016 15:10:32
Number of Replicates: 3
Autosampler Position: 336
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82837.5	1.4				ug/L	74699	Standard
	Be	9	20.0	75.0	**0.0280**	0.015	53.9	ug/L	8	Standard
	Al	27	15988.5	12.5	**0.0848**	0.014	16.0	ug/L	1493	Standard
	Sc	45	64671.1	2.4				ug/L	61029	Standard
	Ti	47	45.0	20.0	**-0.0015**	0.023	1562.0	ug/L	35	Standard
	V	51	1647.0	1.8	**0.0017**	0.004	236.7	ug/L	1435	Standard
	Cr	52	11218.5	4.3	**0.3456**	0.052	15.0	ug/L	7635	Standard
	Cr	53	1063.4	5.0	**0.7116**	0.068	9.6	ug/L	363	Standard
	Mn	55	1756.1	5.1	**0.0801**	0.011	14.2	ug/L	747	Standard
	Co	59	312.3	6.3	**0.0055**	0.003	48.1	ug/L	234	Standard
	Ni	60	1377.4	15.3	**0.5415**	0.099	18.4	ug/L	215	Standard
	Cu	65	261.7	8.1	**0.0235**	0.011	45.2	ug/L	104	Standard
	Zn	66	1613.1	3.7	**0.8814**	0.040	4.6	ug/L	363	Standard
>	**Ge**	72	547319.5	1.5				ug/L	495142	Standard
	As	75	-42.5	100.6	**0.0255**	0.034	133.0	ug/L	-50	Standard
	Se	82	18.0	18.7	**0.0100**	0.027	272.2	ug/L	14	Standard
	Se-1	77	65.0	16.6	**0.1841**	0.145	78.6	ug/L	42	Standard
>	**Ga**	71	66.7	22.9				mg/L	42	Standard
	Rb	85	36.7	56.8				ug/L	17	Standard
	Y	89	450480.7	0.9				ug/L	417098	Standard
>	**Rh**	103	26.7	10.8				ug/L	13	Standard
	Mo	98	139.4	1.7	**0.0435**	0.001	2.2	ug/L	11	Standard
	Ag	107	102.7	18.1	**0.0019**	0.003	140.0	ug/L	68	Standard
	Cd	111	13.1	71.9	**-0.0020**	0.004	209.6	mg/L	2	Standard
	Cd	114	24.5	136.0	**0.0024**	0.006	256.9	ug/L	12	Standard
>	**In**	115	624818.4	0.5				ug/L	587441	Standard
	Sn	118	765.0	2.8	**0.0284**	0.003	12.0	ug/L	497	Standard
	Sb	123	189.1	20.4	**0.0339**	0.008	23.4	ug/L	78	Standard
	Ba	135	83.0	7.9	**0.0221**	0.003	14.0	ug/L	24	Standard
	Ce	140	301.7	18.8				ug/L	25	Standard
>	**Tb**	159	992078.8	0.6				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	40.0	57.5	**0.0027**	0.003	94.2	ug/L	4	Standard
	Tl	205	43.3	29.0	**0.0058**	0.002	28.5	ug/L	2	Standard
	Pb	206	319.7	11.1	**-0.0160**	0.006	36.6	ug/L	240	Standard
	Pb	207	262.0	4.6	**-0.0191**	0.002	12.7	ug/L	198	Standard
	Pb	208	1094.7	2.8	**-0.0202**	0.002	8.1	ug/L	883	Standard
	U	238	22.0	58.0	**-0.0015**	0.002	114.9	ug/L	3	Standard
>	**Bi**	209	568214.1	0.6				ug/L	523984	Standard

Sample ID: PBS M5 WG573099-02
Report Date/Time: Wednesday, June 29, 2016 15:12:48
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	43.3	17.6	**-0.0124**	0.008	67.7	mg/L	37	Standard
K	39	35.0	51.5	**-0.0416**	0.036	86.5	mg/L	38	Standard
Ca	43	30.0	28.9	**-0.7286**	0.543	74.5	mg/L	10	Standard
Fe	54	174.1	18.8	**0.0014**	0.029	2030.6	mg/L	179	Standard
Fe	57	266.7	22.0	**-0.2161**	0.204	94.4	mg/L	335	Standard
> **Sc-1**	45	64671.1	2.4				mg/L	61029	Standard
Cl	35	28617.0	1.0				ug/L	20195	Standard
Kr	83	2.0	100.0				ug/L	4	Standard
Br	81	1953.5	6.4				ug/L	1547	Standard
P	31	19132.1	3.4				ug/L	15286	Standard
S	34	2278.5	6.0				ug/L	1572	Standard
Sr	88	68.3	29.6				ug/L	73	Standard
C	12	83.3	13.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.5	39.0				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	23.3	49.5				mg/L	13	Standard
I	127	8569.1	14.4				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		110.895	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		110.538	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBS M5 WG573099-02
Report Date/Time: Wednesday, June 29, 2016 15:12:48
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.363
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.441
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBS M5 WG573099-02
Report Date/Time: Wednesday, June 29, 2016 15:12:48
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSS M5 WG573099-03
Sample Date/Time: Wednesday, June 29, 2016 15:13:43
Number of Replicates: 3
Autosampler Position: 337
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83551.5	2.3				ug/L	74699	Standard
	Be	9	25751.9	2.9	**26.1593**	0.514	2.0	ug/L	8	Standard
	Al	27	17496.8	6.9	**0.0930**	0.006	6.2	ug/L	1493	Standard
	Sc	45	66115.6	1.8				ug/L	61029	Standard
	Ti	47	52.7	14.3	**0.0175**	0.016	92.6	ug/L	35	Standard
	V	51	220707.2	1.1	**26.4464**	0.390	1.5	ug/L	1435	Standard
	Cr	52	221818.0	0.7	**26.2660**	0.546	2.1	ug/L	7635	Standard
	Cr	53	27359.7	0.9	**26.6591**	0.406	1.5	ug/L	363	Standard
	Mn	55	204818.9	1.5	**26.6717**	0.451	1.7	ug/L	747	Standard
	Co	59	196776.0	1.4	**27.1279**	0.304	1.1	ug/L	234	Standard
	Ni	60	55651.4	0.4	**26.9765**	0.492	1.8	ug/L	215	Standard
	Cu	65	54836.4	0.8	**26.7594**	0.556	2.1	ug/L	104	Standard
	Zn	66	32934.1	1.1	**27.0387**	0.809	3.0	ug/L	363	Standard
>	**Ge**	72	543210.6	2.2				ug/L	495142	Standard
	As	75	31720.5	1.5	**25.8765**	0.351	1.4	ug/L	-50	Standard
	Se	82	3103.9	1.9	**25.6428**	0.383	1.5	ug/L	14	Standard
	Se-1	77	2231.2	2.7	**27.2211**	0.278	1.0	ug/L	42	Standard
>	**Ga**	71	55.0	47.2				mg/L	42	Standard
	Rb	85	35.0	42.9				ug/L	17	Standard
	Y	89	467716.3	0.8				ug/L	417098	Standard
>	**Rh**	103	13.3	43.3				ug/L	13	Standard
	Mo	98	194.8	10.9	**0.0603**	0.008	12.8	ug/L	11	Standard
	Ag	107	10349.2	1.8	**1.4201**	0.038	2.7	ug/L	68	Standard
	Cd	111	56717.0	1.5	**24.4699**	0.504	2.1	mg/L	2	Standard
	Cd	114	136634.8	1.7	**24.0382**	0.506	2.1	ug/L	12	Standard
>	**In**	115	651063.9	0.8				ug/L	587441	Standard
	Sn	118	761.7	5.5	**0.0229**	0.007	28.6	ug/L	497	Standard
	Sb	123	128047.0	1.0	**24.7441**	0.436	1.8	ug/L	78	Standard
	Ba	135	54350.3	0.9	**23.9455**	0.395	1.6	ug/L	24	Standard
	Ce	140	61.7	26.1				ug/L	25	Standard
>	**Tb**	159	1033159.0	0.5				ug/L	939701	Standard
	Ho	165	16.7	17.3				ug/L	10	Standard
	Tl	203	227266.1	0.5	**24.7175**	0.218	0.9	ug/L	4	Standard
	Tl	205	203485.9	1.6	**25.5015**	0.511	2.0	ug/L	2	Standard
	Pb	206	152382.5	0.6	**25.5050**	0.255	1.0	ug/L	240	Standard
	Pb	207	130131.8	0.7	**24.5130**	0.243	1.0	ug/L	198	Standard
	Pb	208	530327.1	0.9	**25.2835**	0.262	1.0	ug/L	883	Standard
	U	238	189200.1	1.2	**24.6397**	0.334	1.4	ug/L	3	Standard
>	**Bi**	209	584433.1	0.4				ug/L	523984	Standard

Sample ID: LCSS M5 WG573099-03
Report Date/Time: Wednesday, June 29, 2016 15:16:00
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	41.7	36.7	**-0.0156**	0.017	112.1	mg/L	37	Standard
K	39	38.3	52.7	**-0.0354**	0.042	119.1	mg/L	38	Standard
Ca	43	25.0	52.9	**-1.0705**	0.758	70.8	mg/L	10	Standard
Fe	54	141.4	21.3	**-0.0274**	0.025	92.5	mg/L	179	Standard
Fe	57	241.7	15.7	**-0.3151**	0.126	40.1	mg/L	335	Standard
Sc-1	45	66115.6	1.8				mg/L	61029	Standard
Cl	35	28385.2	0.6				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	2056.8	1.5				ug/L	1547	Standard
P	31	18911.8	2.1				ug/L	15286	Standard
S	34	2256.8	4.9				ug/L	1572	Standard
Sr	88	70.0	39.8				ug/L	73	Standard
C	12	60.0	66.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	9.5	5.0				mg/L	6	Standard
Ho-1	165	16.7	17.3				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	6019.5	14.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		111.851	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		109.708	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: LCSS M5 WG573099-03
Report Date/Time: Wednesday, June 29, 2016 15:16:00
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.831
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.536
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSS M5 WG573099-03
Report Date/Time: Wednesday, June 29, 2016 15:16:00
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088803 WG573099-01
Sample Date/Time: Wednesday, June 29, 2016 15:16:55
Number of Replicates: 3
Autosampler Position: 338
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87864.2	2.8				ug/L	74699	Standard
	Be	9	1871.8	7.3	**1.8142**	0.100	5.5	ug/L	8	Standard
	Al	27	17464438.5	2.2	**101.0382**	1.686	1.7	ug/L	1493	Standard
	Sc	45	88916.8	1.0				ug/L	61029	Standard
	Ti	47	8105.2	1.6	**20.2503**	0.364	1.8	ug/L	35	Standard
	V	51	266665.9	0.8	**33.4185**	0.188	0.6	ug/L	1435	Standard
	Cr	52	308197.4	0.8	**38.5684**	0.415	1.1	ug/L	7635	Standard
	Cr	53	37658.4	2.2	**38.4602**	0.742	1.9	ug/L	363	Standard
	Mn	55	19789542.3	0.8	**2705.9402**	31.717	1.2	ug/L	747	Standard
	Co	59	276278.5	0.6	**39.7944**	0.401	1.0	ug/L	234	Standard
	Ni	60	160372.4	1.4	**81.4241**	1.275	1.6	ug/L	215	Standard
	Cu	65	107769.0	1.4	**55.0229**	0.902	1.6	ug/L	104	Standard
	Zn	66	213527.8	0.7	**185.6475**	1.406	0.8	ug/L	363	Standard
>	**Ge**	72	520092.2	0.4				ug/L	495142	Standard
	As	75	6170.4	1.0	**5.3042**	0.070	1.3	ug/L	-50	Standard
	Se	82	90.8	2.7	**0.6494**	0.023	3.5	ug/L	14	Standard
	Se-1	77	157.7	4.9	**1.4330**	0.106	7.4	ug/L	42	Standard
>	**Ga**	71	47343.9	1.1				mg/L	42	Standard
	Rb	85	380728.2	1.5				ug/L	17	Standard
	Y	89	1106618.3	0.2				ug/L	417098	Standard
>	**Rh**	103	63.3	24.1				ug/L	13	Standard
	Mo	98	980.4	3.2	**0.3366**	0.010	3.0	ug/L	11	Standard
	Ag	107	955.0	6.8	**0.1236**	0.011	9.1	ug/L	68	Standard
	Cd	111	2225.7	2.0	**0.9829**	0.017	1.8	mg/L	2	Standard
	Cd	114	5479.9	1.8	**0.9924**	0.001	0.1	ug/L	12	Standard
>	**In**	115	631161.9	1.7				ug/L	587441	Standard
	Sn	118	413.3	14.8	**-0.0289**	0.009	31.0	ug/L	497	Standard
	Sb	123	370.7	34.4	**0.0695**	0.025	35.4	ug/L	78	Standard
	Ba	135	385428.4	1.0	**175.2729**	2.367	1.4	ug/L	24	Standard
	Ce	140	2007459.4	3.6				ug/L	25	Standard
>	**Tb**	159	1048778.2	1.5				ug/L	939701	Standard
	Ho	165	53168.1	1.9				ug/L	10	Standard
	Tl	203	2332.8	1.5	**0.2668**	0.007	2.6	ug/L	4	Standard
	Tl	205	2056.8	1.1	**0.2730**	0.001	0.5	ug/L	2	Standard
	Pb	206	102474.4	0.3	**18.1326**	0.189	1.0	ug/L	240	Standard
	Pb	207	82544.6	1.5	**16.4326**	0.141	0.9	ug/L	198	Standard
	Pb	208	353310.3	0.4	**17.8060**	0.180	1.0	ug/L	883	Standard
	U	238	9355.9	0.7	**1.2854**	0.017	1.3	ug/L	3	Standard
>	**Bi**	209	552227.6	1.4				ug/L	523984	Standard

Sample ID: L1606088803 WG573099-01
Report Date/Time: Wednesday, June 29, 2016 15:19:11
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**13.3403**	23.097	173.1	mg/L	0	Standard
Mg	24	320.0	5.6	**0.2185**	0.016	7.5	mg/L	37	Standard
K	39	1080.0	10.5	**1.5233**	0.156	10.2	mg/L	38	Standard
Ca	43	916.7	12.0	**37.2112**	4.625	12.4	mg/L	10	Standard
Fe	54	75887.5	2.1	**44.0256**	1.050	2.4	mg/L	179	Standard
Fe	57	19502.5	1.2	**44.4256**	0.652	1.5	mg/L	335	Standard
> **Sc-1**	45	88916.8	1.0				mg/L	61029	Standard
Cl	35	27513.6	0.5				ug/L	20195	Standard
Kr	83	0.3	173.2				ug/L	4	Standard
Br	81	3467.1	2.9				ug/L	1547	Standard
P	31	21865.8	2.4				ug/L	15286	Standard
S	34	2631.9	4.0				ug/L	1572	Standard
Sr	88	220.0	6.0				ug/L	73	Standard
C	12	66.7	37.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	36.7	15.7				mg/L	3	Standard
Dy	164	86599.4	2.5				mg/L	6	Standard
Ho-1	165	53168.1	1.9				mg/L	10	Standard
Er	166	47260.3	0.7				mg/L	13	Standard
I	127	6809.9	9.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		117.625	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		105.039	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606088803 WG573099-01
Report Date/Time: Wednesday, June 29, 2016 15:19:11
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	107.443
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	105.390
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606088803 WG573099-01
Report Date/Time: Wednesday, June 29, 2016 15:19:11
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088803S WG573099-04
Sample Date/Time: Wednesday, June 29, 2016 15:20:06
Number of Replicates: 3
Autosampler Position: 339
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88784.5	2.9				ug/L	74699	Standard
	Be	9	27508.3	0.4	**26.3080**	0.631	2.4	ug/L	8	Standard
	Al	27	20794150.5	3.6	**119.0173**	0.902	0.8	ug/L	1493	Standard
	Sc	45	96992.9	2.3				ug/L	61029	Standard
	Ti	47	11606.4	2.4	**28.2553**	0.997	3.5	ug/L	35	Standard
	V	51	594819.7	3.2	**72.7593**	3.186	4.4	ug/L	1435	Standard
	Cr	52	602098.1	1.8	**74.2282**	2.325	3.1	ug/L	7635	Standard
	Cr	53	74295.1	1.5	**74.1206**	1.783	2.4	ug/L	363	Standard
	Mn	55	28090211.5	1.2	**3736.5671**	85.696	2.3	ug/L	747	Standard
	Co	59	552502.8	1.4	**77.4533**	1.929	2.5	ug/L	234	Standard
	Ni	60	270132.5	1.6	**133.5117**	3.898	2.9	ug/L	215	Standard
	Cu	65	201432.2	2.2	**100.1412**	3.335	3.3	ug/L	104	Standard
	Zn	66	320182.0	2.8	**271.0655**	11.001	4.1	ug/L	363	Standard
>	**Ge**	72	534719.5	1.4				ug/L	495142	Standard
	As	75	41934.7	2.3	**34.7370**	1.197	3.4	ug/L	-50	Standard
	Se	82	2480.2	1.5	**20.7934**	0.607	2.9	ug/L	14	Standard
	Se-1	77	1799.1	2.3	**22.1951**	0.836	3.8	ug/L	42	Standard
>	**Ga**	71	63861.1	4.0				mg/L	42	Standard
	Rb	85	554637.9	1.8				ug/L	17	Standard
	Y	89	1250941.0	3.4				ug/L	417098	Standard
>	**Rh**	103	88.3	8.6				ug/L	13	Standard
	Mo	98	1454.1	4.8	**0.5217**	0.033	6.4	ug/L	11	Standard
	Ag	107	4814.4	4.1	**0.7006**	0.018	2.6	ug/L	68	Standard
	Cd	111	56224.0	1.3	**25.9728**	0.111	0.4	mg/L	2	Standard
	Cd	114	138439.7	3.6	**26.0717**	0.493	1.9	ug/L	12	Standard
>	**In**	115	608045.3	1.7				ug/L	587441	Standard
	Sn	118	518.3	2.0	**-0.0090**	0.001	16.4	ug/L	497	Standard
	Sb	123	2325.7	1.7	**0.4771**	0.008	1.7	ug/L	78	Standard
	Ba	135	587827.7	1.5	**277.4602**	0.795	0.3	ug/L	24	Standard
	Ce	140	2927642.8	0.6				ug/L	25	Standard
>	**Tb**	159	1033681.4	2.7				ug/L	939701	Standard
	Ho	165	68191.4	0.6				ug/L	10	Standard
	Tl	203	212892.3	0.8	**25.4932**	0.083	0.3	ug/L	4	Standard
	Tl	205	191048.7	1.2	**26.3598**	0.091	0.3	ug/L	2	Standard
	Pb	206	260602.6	2.5	**48.0843**	0.971	2.0	ug/L	240	Standard
	Pb	207	216580.3	1.8	**44.9759**	0.652	1.4	ug/L	198	Standard
	Pb	208	918219.8	1.2	**48.2652**	0.290	0.6	ug/L	883	Standard
	U	238	196774.7	2.4	**28.2131**	0.460	1.6	ug/L	3	Standard
>	**Bi**	209	530811.0	1.1				ug/L	523984	Standard

Sample ID: L1606088803S WG573099-04
Report Date/Time: Wednesday, June 29, 2016 15:22:23
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**12.2205**	21.158	173.1	mg/L	0	Standard
Mg	24	365.0	3.6	**0.2314**	0.007	3.2	mg/L	37	Standard
K	39	1516.7	5.6	**1.9963**	0.137	6.9	mg/L	38	Standard
Ca	43	726.7	5.8	**26.3770**	1.992	7.6	mg/L	10	Standard
Fe	54	108682.6	3.2	**57.8341**	0.799	1.4	mg/L	179	Standard
Fe	57	27718.7	1.4	**58.2169**	0.892	1.5	mg/L	335	Standard
Sc-1	45	96992.9	2.3				mg/L	61029	Standard
Cl	35	26216.7	4.4				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	4344.0	8.1				ug/L	1547	Standard
P	31	22273.0	3.2				ug/L	15286	Standard
S	34	2523.5	5.1				ug/L	1572	Standard
Sr	88	263.3	11.4				ug/L	73	Standard
C	12	56.7	10.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	52.9				mg/L	3	Standard
Dy	164	111985.6	3.7				mg/L	6	Standard
Ho-1	165	68191.4	0.6				mg/L	10	Standard
Er	166	57569.3	4.1				mg/L	13	Standard
I	127	6513.1	8.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		118.857	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		107.993	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606088803S WG573099-04
Report Date/Time: Wednesday, June 29, 2016 15:22:23
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.507
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.303
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ni 60 Upper, S, EEE	Ni	60	

Sample ID: L1606088803S WG573099-04
Report Date/Time: Wednesday, June 29, 2016 15:22:23
Page 3

Approved: July 04, 2016

Cu 65 Upper, S, EEE	Cu	65
Zn 66 Upper, S, EEE	Zn	66
Ba 135 Upper, S, EEE	Ba	135

Sample ID: L1606088803S WG573099-04
Report Date/Time: Wednesday, June 29, 2016 15:22:23
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088803SD WG573099-05
Sample Date/Time: Wednesday, June 29, 2016 15:23:18
Number of Replicates: 3
Autosampler Position: 340
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86779.6	2.9				ug/L	74699	Standard
	Be	9	26783.7	4.5	**26.1889**	0.607	2.3	ug/L	8	Standard
	Al	27	17495490.1	2.1	**102.4852**	1.702	1.7	ug/L	1493	Standard
	Sc	45	84971.1	3.4				ug/L	61029	Standard
	Ti	47	11032.7	2.5	**25.6633**	0.769	3.0	ug/L	35	Standard
	V	51	492228.0	0.7	**57.5050**	0.764	1.3	ug/L	1435	Standard
	Cr	52	556079.4	1.4	**65.4064**	1.154	1.8	ug/L	7635	Standard
	Cr	53	67044.7	1.2	**63.8916**	0.827	1.3	ug/L	363	Standard
	Mn	55	16288704.5	1.7	**2071.2097**	47.470	2.3	ug/L	747	Standard
	Co	59	487642.9	2.1	**65.3439**	1.764	2.7	ug/L	234	Standard
	Ni	60	230286.3	2.0	**108.7653**	2.155	2.0	ug/L	215	Standard
	Cu	65	186555.0	1.4	**88.6352**	1.512	1.7	ug/L	104	Standard
	Zn	66	269449.8	1.5	**217.9309**	3.745	1.7	ug/L	363	Standard
>	**Ge**	72	559301.6	0.7				ug/L	495142	Standard
	As	75	32569.4	2.0	**25.8033**	0.648	2.5	ug/L	-50	Standard
	Se	82	2764.3	2.2	**22.1623**	0.630	2.8	ug/L	14	Standard
	Se-1	77	1950.8	4.1	**23.0256**	1.042	4.5	ug/L	42	Standard
>	**Ga**	71	55862.8	2.2				mg/L	42	Standard
	Rb	85	472783.4	2.1				ug/L	17	Standard
	Y	89	931018.8	1.5				ug/L	417098	Standard
>	**Rh**	103	36.7	47.9				ug/L	13	Standard
	Mo	98	833.5	3.5	**0.3003**	0.015	5.0	ug/L	11	Standard
	Ag	107	3972.2	0.3	**0.5833**	0.011	1.9	ug/L	68	Standard
	Cd	111	54996.2	0.6	**25.7204**	0.316	1.2	mg/L	2	Standard
	Cd	114	133681.6	2.4	**25.4888**	0.206	0.8	ug/L	12	Standard
>	**In**	115	600647.8	1.6				ug/L	587441	Standard
	Sn	118	466.7	16.3	**-0.0167**	0.012	70.0	ug/L	497	Standard
	Sb	123	3387.1	2.7	**0.7053**	0.008	1.2	ug/L	78	Standard
	Ba	135	467382.4	1.0	**223.3343**	2.175	1.0	ug/L	24	Standard
	Ce	140	1516729.6	1.2				ug/L	25	Standard
>	**Tb**	159	1007426.2	2.1				ug/L	939701	Standard
	Ho	165	39640.1	1.9				ug/L	10	Standard
	Tl	203	211811.1	1.3	**24.5598**	0.357	1.5	ug/L	4	Standard
	Tl	205	187359.9	1.8	**25.0339**	0.597	2.4	ug/L	2	Standard
	Pb	206	219015.7	1.1	**39.1225**	0.790	2.0	ug/L	240	Standard
	Pb	207	182856.8	0.5	**36.7608**	0.679	1.8	ug/L	198	Standard
	Pb	208	768036.7	1.8	**39.0813**	0.960	2.5	ug/L	883	Standard
	U	238	187429.9	1.2	**26.0258**	0.546	2.1	ug/L	3	Standard
>	**Bi**	209	548237.3	1.8				ug/L	523984	Standard

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 1

Approved: July 04, 2016

Na	23	6.7	43.3	**55.5842**	25.460	45.8	mg/L	0	Standard
Mg	24	348.3	15.5	**0.2571**	0.039	15.2	mg/L	37	Standard
K	39	1421.7	3.5	**2.1441**	0.098	4.5	mg/L	38	Standard
Ca	43	383.3	17.4	**14.9477**	3.560	23.8	mg/L	10	Standard
Fe	54	86673.1	2.4	**52.6645**	1.510	2.9	mg/L	179	Standard
Fe	57	21795.7	3.1	**52.2085**	3.369	6.5	mg/L	335	Standard
> **Sc-1**	45	84971.1	3.4				mg/L	61029	Standard
Cl	35	24892.4	3.3				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	3257.0	4.5				ug/L	1547	Standard
P	31	20836.0	2.5				ug/L	15286	Standard
S	34	2406.9	4.4				ug/L	1572	Standard
Sr	88	191.7	46.0				ug/L	73	Standard
C	12	76.7	15.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	20.0				mg/L	3	Standard
Dy	164	63526.1	1.7				mg/L	6	Standard
Ho-1	165	39640.1	1.9				mg/L	10	Standard
Er	166	35224.2	2.0				mg/L	13	Standard
I	127	5139.2	9.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		116.173	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		112.958	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	102.248
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.629
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ni 60 Upper, S, EEE	Ni	60	

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 3

Approved: July 04, 2016

Zn 66 Upper, S, EEE	Zn	66
Ba 135 Upper, S, EEE	Ba	135

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088301
Sample Date/Time: Wednesday, June 29, 2016 15:26:29
Number of Replicates: 3
Autosampler Position: 341
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84384.4	1.7				ug/L	74699	Standard
	Be	9	1073.4	1.5	**1.0872**	0.032	3.0	ug/L	8	Standard
	Al	27	11940771.9	2.8	**71.9363**	2.792	3.9	ug/L	1493	Standard
	Sc	45	76626.6	1.2				ug/L	61029	Standard
	Ti	47	4374.3	2.0	**10.2834**	0.247	2.4	ug/L	35	Standard
	V	51	162703.8	1.0	**19.2014**	0.098	0.5	ug/L	1435	Standard
	Cr	52	132016.8	1.7	**15.0167**	0.347	2.3	ug/L	7635	Standard
	Cr	53	16153.6	0.8	**15.4073**	0.063	0.4	ug/L	363	Standard
	Mn	55	3262876.5	0.6	**421.8414**	5.005	1.2	ug/L	747	Standard
	Co	59	67832.8	0.7	**9.2120**	0.089	1.0	ug/L	234	Standard
	Ni	60	42623.2	1.3	**20.3743**	0.274	1.3	ug/L	215	Standard
	Cu	65	31205.8	0.3	**14.9935**	0.149	1.0	ug/L	104	Standard
	Zn	66	53134.7	1.9	**43.3429**	0.711	1.6	ug/L	363	Standard
>	**Ge**	72	549915.9	0.8				ug/L	495142	Standard
	As	75	3335.6	3.0	**2.7409**	0.080	2.9	ug/L	-50	Standard
	Se	82	46.6	19.0	**0.2449**	0.076	30.9	ug/L	14	Standard
	Se-1	77	111.7	5.1	**0.7542**	0.066	8.8	ug/L	42	Standard
>	**Ga**	71	17852.2	2.1				mg/L	42	Standard
	Rb	85	173572.1	2.0				ug/L	17	Standard
	Y	89	703667.4	2.2				ug/L	417098	Standard
>	**Rh**	103	33.3	34.6				ug/L	13	Standard
	Mo	98	357.4	3.8	**0.1158**	0.007	6.2	ug/L	11	Standard
	Ag	107	550.0	6.6	**0.0636**	0.007	11.1	ug/L	68	Standard
	Cd	111	273.3	3.2	**0.1105**	0.006	5.2	mg/L	2	Standard
	Cd	114	617.1	4.0	**0.1069**	0.006	5.7	ug/L	12	Standard
>	**In**	115	648944.8	3.3				ug/L	587441	Standard
	Sn	118	606.7	6.6	**-0.0008**	0.003	418.3	ug/L	497	Standard
	Sb	123	137.4	14.1	**0.0224**	0.003	13.8	ug/L	78	Standard
	Ba	135	318337.1	2.2	**140.8176**	2.386	1.7	ug/L	24	Standard
	Ce	140	1211648.5	1.9				ug/L	25	Standard
>	**Tb**	159	1041543.0	2.5				ug/L	939701	Standard
	Ho	165	19666.1	3.2				ug/L	10	Standard
	Tl	203	1035.0	3.0	**0.1143**	0.004	3.7	ug/L	4	Standard
	Tl	205	953.4	8.0	**0.1233**	0.009	7.5	ug/L	2	Standard
	Pb	206	127279.7	1.1	**21.9443**	0.442	2.0	ug/L	240	Standard
	Pb	207	103046.6	1.5	**19.9907**	0.426	2.1	ug/L	198	Standard
	Pb	208	441148.5	1.4	**21.6612**	0.196	0.9	ug/L	883	Standard
	U	238	5034.2	0.5	**0.6713**	0.005	0.8	ug/L	3	Standard
>	**Bi**	209	567168.2	1.1				ug/L	523984	Standard

Sample ID: L1606088301
Report Date/Time: Wednesday, June 29, 2016 15:28:46
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	340.0	3.9	**0.2844**	0.011	3.7	mg/L	37	Standard
K	39	688.3	2.3	**1.0975**	0.019	1.8	mg/L	38	Standard
Ca	43	145.0	15.8	**4.7718**	1.160	24.3	mg/L	10	Standard
Fe	54	17324.9	3.4	**11.5598**	0.337	2.9	mg/L	179	Standard
Fe	57	4550.7	3.6	**11.2444**	0.399	3.5	mg/L	335	Standard
Sc-1	45	76626.6	1.2				mg/L	61029	Standard
Cl	35	25558.2	1.3				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	2263.5	8.1				ug/L	1547	Standard
P	31	20323.6	0.9				ug/L	15286	Standard
S	34	2258.5	1.1				ug/L	1572	Standard
Sr	88	101.7	10.2				ug/L	73	Standard
C	12	116.7	13.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	49.5				mg/L	3	Standard
Dy	164	32996.6	2.0				mg/L	6	Standard
Ho-1	165	19666.1	3.2				mg/L	10	Standard
Er	166	17349.9	3.1				mg/L	13	Standard
I	127	11507.7	1.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		112.966	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		111.062	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606088301
Report Date/Time: Wednesday, June 29, 2016 15:28:46
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.470
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.241
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606088301
Report Date/Time: Wednesday, June 29, 2016 15:28:46
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201
Sample Date/Time: Wednesday, June 29, 2016 15:29:41
Number of Replicates: 3
Autosampler Position: 342
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84277.1	1.0				ug/L	74699	Standard
	Be	9	1593.4	6.0	**1.6114**	0.083	5.1	ug/L	8	Standard
	Al	27	15243679.1	1.9	**91.9299**	1.967	2.1	ug/L	1493	Standard
	Sc	45	83538.1	3.1				ug/L	61029	Standard
	Ti	47	10307.8	1.1	**23.4331**	0.360	1.5	ug/L	35	Standard
	V	51	410563.1	0.8	**46.8597**	0.438	0.9	ug/L	1435	Standard
	Cr	52	261972.4	0.3	**29.5747**	0.180	0.6	ug/L	7635	Standard
	Cr	53	32057.2	1.7	**29.6933**	0.417	1.4	ug/L	363	Standard
	Mn	55	12908980.1	3.2	**1604.9549**	58.354	3.6	ug/L	747	Standard
	Co	59	181241.6	3.1	**23.7226**	0.846	3.6	ug/L	234	Standard
	Ni	60	66685.8	0.9	**30.7063**	0.259	0.8	ug/L	215	Standard
	Cu	65	73374.2	0.7	**34.0220**	0.398	1.2	ug/L	104	Standard
	Zn	66	143517.2	0.9	**113.2755**	1.322	1.2	ug/L	363	Standard
>	**Ge**	72	572023.7	0.5				ug/L	495142	Standard
	As	75	16668.5	2.4	**12.9416**	0.353	2.7	ug/L	-50	Standard
	Se	82	187.7	6.4	**1.3421**	0.095	7.1	ug/L	14	Standard
	Se-1	77	232.0	7.5	**2.1278**	0.217	10.2	ug/L	42	Standard
>	**Ga**	71	37605.0	3.4				mg/L	42	Standard
	Rb	85	431124.7	1.0				ug/L	17	Standard
	Y	89	907846.1	0.7				ug/L	417098	Standard
>	**Rh**	103	28.3	27.0				ug/L	13	Standard
	Mo	98	1317.8	2.4	**0.4515**	0.009	1.9	ug/L	11	Standard
	Ag	107	1338.4	1.8	**0.1770**	0.004	2.5	ug/L	68	Standard
	Cd	111	803.3	6.8	**0.3473**	0.022	6.2	mg/L	2	Standard
	Cd	114	1888.8	7.2	**0.3384**	0.023	6.7	ug/L	12	Standard
>	**In**	115	635240.6	1.1				ug/L	587441	Standard
	Sn	118	608.3	2.5	**0.0016**	0.002	109.2	ug/L	497	Standard
	Sb	123	242.0	3.6	**0.0438**	0.002	4.9	ug/L	78	Standard
	Ba	135	645480.7	0.7	**291.6323**	1.817	0.6	ug/L	24	Standard
	Ce	140	2459220.7	0.8				ug/L	25	Standard
>	**Tb**	159	1043691.8	1.6				ug/L	939701	Standard
	Ho	165	34592.8	2.2				ug/L	10	Standard
	Tl	203	3164.7	1.5	**0.3540**	0.007	2.0	ug/L	4	Standard
	Tl	205	2843.6	5.4	**0.3684**	0.020	5.4	ug/L	2	Standard
	Pb	206	325793.2	0.7	**56.4391**	0.926	1.6	ug/L	240	Standard
	Pb	207	262124.0	0.3	**51.1005**	0.344	0.7	ug/L	198	Standard
	Pb	208	1122453.0	1.2	**55.3927**	1.161	2.1	ug/L	883	Standard
	U	238	18503.6	0.6	**2.4863**	0.037	1.5	ug/L	3	Standard
>	**Bi**	209	565563.6	0.9				ug/L	523984	Standard

Sample ID: L1606091201
Report Date/Time: Wednesday, June 29, 2016 15:31:58
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	148.3	19.2	**0.0751**	0.029	38.8	mg/L	37	Standard
K	39	1071.7	2.8	**1.6167**	0.038	2.4	mg/L	38	Standard
Ca	43	301.7	16.4	**11.4520**	2.661	23.2	mg/L	10	Standard
Fe	54	45977.7	1.1	**28.3556**	0.839	3.0	mg/L	179	Standard
Fe	57	11742.9	4.0	**28.0839**	0.746	2.7	mg/L	335	Standard
> **Sc-1**	45	83538.1	3.1				mg/L	61029	Standard
Cl	35	25586.9	0.8				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	3940.5	1.8				ug/L	1547	Standard
P	31	20170.1	2.1				ug/L	15286	Standard
S	34	2405.2	3.2				ug/L	1572	Standard
Sr	88	111.7	6.8				ug/L	73	Standard
C	12	316.7	15.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	146.7	35.0				mg/L	3	Standard
Dy	164	57194.9	2.2				mg/L	6	Standard
Ho-1	165	34592.8	2.2				mg/L	10	Standard
Er	166	32892.3	1.1				mg/L	13	Standard
I	127	48256.8	4.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		112.823	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.527	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091201
Report Date/Time: Wednesday, June 29, 2016 15:31:58
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	108.137
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	107.935
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606091201
Report Date/Time: Wednesday, June 29, 2016 15:31:58
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201PS WG573717-03
Sample Date/Time: Wednesday, June 29, 2016 15:32:52
Number of Replicates: 3
Autosampler Position: 343
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84925.7	1.3				ug/L	74699	Standard
	Be	9	54409.2	0.5	**54.3741**	0.919	1.7	ug/L	8	Standard
	Al	27	15619580.7	1.4	**93.4869**	2.147	2.3	ug/L	1493	Standard
	Sc	45	84914.0	2.3				ug/L	61029	Standard
	Ti	47	10489.6	2.4	**24.7035**	0.475	1.9	ug/L	35	Standard
	V	51	847646.1	0.7	**100.4344**	1.074	1.1	ug/L	1435	Standard
	Cr	52	696222.3	1.3	**83.2031**	0.580	0.7	ug/L	7635	Standard
	Cr	53	86425.9	2.6	**83.5032**	1.546	1.9	ug/L	363	Standard
	Mn	55	12964778.0	1.1	**1669.4256**	16.996	1.0	ug/L	747	Standard
	Co	59	568224.0	0.4	**77.1151**	0.744	1.0	ug/L	234	Standard
	Ni	60	176707.0	1.9	**84.4902**	0.918	1.1	ug/L	215	Standard
	Cu	65	181984.3	1.8	**87.5571**	0.864	1.0	ug/L	104	Standard
	Zn	66	208968.8	1.8	**171.0579**	1.957	1.1	ug/L	363	Standard
>	**Ge**	72	552256.5	0.8				ug/L	495142	Standard
	As	75	82371.4	1.6	**65.9929**	0.953	1.4	ug/L	-50	Standard
	Se	82	6597.3	1.6	**53.7542**	0.590	1.1	ug/L	14	Standard
	Se-1	77	4584.4	1.6	**55.6563**	1.034	1.9	ug/L	42	Standard
>	**Ga**	71	37593.3	1.5				mg/L	42	Standard
	Rb	85	428376.8	2.4				ug/L	17	Standard
	Y	89	890825.8	0.7				ug/L	417098	Standard
>	**Rh**	103	50.0	10.0				ug/L	13	Standard
	Mo	98	1335.6	2.9	**0.4503**	0.014	3.2	ug/L	11	Standard
	Ag	107	364007.1	1.3	**50.8126**	0.330	0.6	ug/L	68	Standard
	Cd	111	115695.5	1.2	**50.3468**	0.384	0.8	mg/L	2	Standard
	Cd	114	280904.1	1.9	**49.8365**	0.143	0.3	ug/L	12	Standard
>	**In**	115	645589.3	1.8				ug/L	587441	Standard
	Sn	118	755.0	4.3	**0.0229**	0.003	13.2	ug/L	497	Standard
	Sb	123	254065.1	1.3	**49.5173**	0.623	1.3	ug/L	78	Standard
	Ba	135	752068.0	1.2	**334.3566**	2.055	0.6	ug/L	24	Standard
	Ce	140	2468196.2	1.3				ug/L	25	Standard
>	**Tb**	159	1069312.5	1.3				ug/L	939701	Standard
	Ho	165	35102.3	1.1				ug/L	10	Standard
	Tl	203	458369.6	1.1	**50.4504**	0.349	0.7	ug/L	4	Standard
	Tl	205	409805.1	0.9	**51.9806**	1.259	2.4	ug/L	2	Standard
	Pb	206	619020.9	0.8	**105.0729**	1.076	1.0	ug/L	240	Standard
	Pb	207	511972.0	0.6	**97.8107**	1.381	1.4	ug/L	198	Standard
	Pb	208	2150459.1	0.4	**103.9876**	1.211	1.2	ug/L	883	Standard
	U	238	395808.0	0.4	**52.1735**	0.789	1.5	ug/L	3	Standard
>	**Bi**	209	577539.6	1.5				ug/L	523984	Standard

Sample ID: L1606091201PS WG573717-03
Report Date/Time: Wednesday, June 29, 2016 15:35:09
Page 1

Approved: July 04, 2016

Na	23	6.7	43.3	**54.9152**	22.487	40.9	mg/L	0	Standard
Mg	24	118.3	17.1	**0.0444**	0.016	36.3	mg/L	37	Standard
K	39	1026.7	7.6	**1.5167**	0.116	7.6	mg/L	38	Standard
Ca	43	296.7	14.3	**10.9789**	2.237	20.4	mg/L	10	Standard
Fe	54	45524.1	0.4	**27.6104**	0.565	2.0	mg/L	179	Standard
Fe	57	12291.7	1.9	**28.9717**	1.259	4.3	mg/L	335	Standard
> **Sc-1**	45	84914.0	2.3				mg/L	61029	Standard
Cl	35	25819.3	0.8				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	3963.9	6.7				ug/L	1547	Standard
P	31	20629.0	3.4				ug/L	15286	Standard
S	34	2280.2	5.7				ug/L	1572	Standard
Sr	88	103.3	12.2				ug/L	73	Standard
C	12	333.3	17.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	140.0	7.1				mg/L	3	Standard
Dy	164	56592.8	3.3				mg/L	6	Standard
Ho-1	165	35102.3	1.1				mg/L	10	Standard
Er	166	32397.9	1.1				mg/L	13	Standard
I	127	48815.3	4.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.691	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.535	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091201PS WG573717-03
Report Date/Time: Wednesday, June 29, 2016 15:35:09
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.899
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.221
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
V 51 Upper, S, EEE	V	51	
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606091201PS WG573717-03
Report Date/Time: Wednesday, June 29, 2016 15:35:09
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135
Pb 206 Upper, S, EEE	Pb	206
Pb 208 Upper, S, EEE	Pb	208

Sample ID: L1606091201PS WG573717-03
Report Date/Time: Wednesday, June 29, 2016 15:35:09
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201SDL WG573717-04
Sample Date/Time: Wednesday, June 29, 2016 15:36:04
Number of Replicates: 3
Autosampler Position: 344
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83450.9	2.1				ug/L	74699	Standard
	Be	9	378.3	16.7	**0.3916**	0.058	14.9	ug/L	8	Standard
	Al	27	3082760.3	2.5	**18.7625**	0.099	0.5	ug/L	1493	Standard
	Sc	45	72303.8	2.1				ug/L	61029	Standard
	Ti	47	2033.1	4.6	**4.7237**	0.243	5.1	ug/L	35	Standard
	V	51	79797.1	0.9	**9.3216**	0.052	0.6	ug/L	1435	Standard
	Cr	52	57744.9	2.6	**5.9945**	0.158	2.6	ug/L	7635	Standard
	Cr	53	6483.1	2.8	**5.9884**	0.155	2.6	ug/L	363	Standard
	Mn	55	2512242.8	1.3	**324.8710**	1.825	0.6	ug/L	747	Standard
	Co	59	34712.7	2.2	**4.6975**	0.071	1.5	ug/L	234	Standard
	Ni	60	13460.0	2.9	**6.3513**	0.158	2.5	ug/L	215	Standard
	Cu	65	15009.4	2.6	**7.1598**	0.141	2.0	ug/L	104	Standard
	Zn	66	29256.2	2.4	**23.6692**	0.389	1.6	ug/L	363	Standard
>	**Ge**	72	549686.2	0.8				ug/L	495142	Standard
	As	75	3156.6	2.0	**2.5979**	0.033	1.3	ug/L	-50	Standard
	Se	82	47.1	17.6	**0.2490**	0.070	28.2	ug/L	14	Standard
	Se-1	77	77.0	11.9	**0.3268**	0.106	32.4	ug/L	42	Standard
>	**Ga**	71	7136.7	1.1				mg/L	42	Standard
	Rb	85	84221.8	1.7				ug/L	17	Standard
	Y	89	569587.4	0.8				ug/L	417098	Standard
>	**Rh**	103	25.0	91.7				ug/L	13	Standard
	Mo	98	271.1	6.9	**0.0847**	0.006	6.6	ug/L	11	Standard
	Ag	107	378.0	10.8	**0.0387**	0.005	13.9	ug/L	68	Standard
	Cd	111	185.8	10.3	**0.0712**	0.009	12.3	mg/L	2	Standard
	Cd	114	406.5	6.5	**0.0684**	0.004	6.4	ug/L	12	Standard
>	**In**	115	660230.6	1.0				ug/L	587441	Standard
	Sn	118	501.7	8.5	**-0.0183**	0.006	32.3	ug/L	497	Standard
	Sb	123	1797.9	20.8	**0.3384**	0.070	20.8	ug/L	78	Standard
	Ba	135	129725.4	1.6	**56.3802**	0.972	1.7	ug/L	24	Standard
	Ce	140	480517.5	2.7				ug/L	25	Standard
>	**Tb**	159	1042248.3	0.8				ug/L	939701	Standard
	Ho	165	6823.2	4.6				ug/L	10	Standard
	Tl	203	727.7	7.7	**0.0794**	0.006	7.1	ug/L	4	Standard
	Tl	205	616.7	4.2	**0.0795**	0.002	2.4	ug/L	2	Standard
	Pb	206	64359.4	2.2	**11.0114**	0.128	1.2	ug/L	240	Standard
	Pb	207	51665.3	2.6	**9.9429**	0.134	1.3	ug/L	198	Standard
	Pb	208	222946.4	0.6	**10.8660**	0.247	2.3	ug/L	883	Standard
	U	238	3701.1	2.3	**0.4902**	0.006	1.1	ug/L	3	Standard
>	**Bi**	209	569638.6	2.0				ug/L	523984	Standard

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:38:21
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	61.7	28.5	**0.0019**	0.018	942.5	mg/L	37	Standard
K	39	203.3	18.5	**0.2649**	0.067	25.4	mg/L	38	Standard
Ca	43	75.0	26.7	**1.4632**	0.987	67.5	mg/L	10	Standard
Fe	54	9325.9	3.5	**6.5358**	0.204	3.1	mg/L	179	Standard
Fe	57	2630.2	2.1	**6.4718**	0.055	0.8	mg/L	335	Standard
> **Sc-1**	45	72303.8	2.1				mg/L	61029	Standard
Cl	35	24972.6	0.9				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2286.8	8.4				ug/L	1547	Standard
P	31	19043.6	4.4				ug/L	15286	Standard
S	34	2180.2	1.2				ug/L	1572	Standard
Sr	88	81.7	35.9				ug/L	73	Standard
C	12	93.3	22.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	57.3				mg/L	3	Standard
Dy	164	11677.0	2.4				mg/L	6	Standard
Ho-1	165	6823.2	4.6				mg/L	10	Standard
Er	166	6424.7	2.9				mg/L	13	Standard
I	127	12089.8	7.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.717	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.016	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:38:21
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	112.391
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.713
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:38:21
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201SDL WG573717-04
Sample Date/Time: Wednesday, June 29, 2016 15:39:15
Number of Replicates: 3
Autosampler Position: 345
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80082.8	1.5				ug/L	74699	Standard
	Be	9	80.0	10.8	**0.0924**	0.009	10.2	ug/L	8	Standard
	Al	27	638793.5	0.9	**4.0416**	0.057	1.4	ug/L	1493	Standard
	Sc	45	66977.8	2.5				ug/L	61029	Standard
	Ti	47	422.3	5.3	**0.8899**	0.044	4.9	ug/L	35	Standard
	V	51	17807.2	0.8	**1.9190**	0.038	2.0	ug/L	1435	Standard
	Cr	52	18051.7	1.5	**1.1605**	0.015	1.3	ug/L	7635	Standard
	Cr	53	1846.8	4.0	**1.4619**	0.067	4.6	ug/L	363	Standard
	Mn	55	540671.4	1.9	**69.4959**	1.549	2.2	ug/L	747	Standard
	Co	59	7520.2	3.3	**0.9841**	0.038	3.8	ug/L	234	Standard
	Ni	60	2912.3	1.6	**1.2710**	0.019	1.5	ug/L	215	Standard
	Cu	65	3203.0	1.2	**1.4395**	0.003	0.2	ug/L	104	Standard
	Zn	66	8550.1	1.3	**6.5648**	0.083	1.3	ug/L	363	Standard
>	**Ge**	72	552153.4	1.0				ug/L	495142	Standard
	As	75	624.0	5.3	**0.5591**	0.024	4.4	ug/L	-50	Standard
	Se	82	21.8	25.1	**0.0396**	0.043	108.7	ug/L	14	Standard
	Se-1	77	57.0	8.0	**0.0772**	0.050	64.2	ug/L	42	Standard
>	**Ga**	71	1470.1	2.9				mg/L	42	Standard
	Rb	85	17425.0	3.8				ug/L	17	Standard
	Y	89	497203.9	1.3				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	65.7	10.1	**0.0165**	0.003	15.8	ug/L	11	Standard
	Ag	107	156.7	8.7	**0.0087**	0.002	18.0	ug/L	68	Standard
	Cd	111	38.9	11.8	**0.0089**	0.002	23.3	mg/L	2	Standard
	Cd	114	108.8	14.4	**0.0170**	0.003	15.1	ug/L	12	Standard
>	**In**	115	651862.8	1.7				ug/L	587441	Standard
	Sn	118	480.0	4.2	**-0.0206**	0.003	16.7	ug/L	497	Standard
	Sb	123	490.3	31.4	**0.0908**	0.031	34.3	ug/L	78	Standard
	Ba	135	26415.7	0.1	**11.6170**	0.204	1.8	ug/L	24	Standard
	Ce	140	98329.7	0.4				ug/L	25	Standard
>	**Tb**	159	1007324.2	1.5				ug/L	939701	Standard
	Ho	165	1405.1	5.7				ug/L	10	Standard
	Tl	203	169.0	1.6	**0.0177**	0.000	0.8	ug/L	4	Standard
	Tl	205	175.0	8.6	**0.0234**	0.002	8.3	ug/L	2	Standard
	Pb	206	13505.0	1.4	**2.3274**	0.057	2.5	ug/L	240	Standard
	Pb	207	10900.6	1.3	**2.1089**	0.044	2.1	ug/L	198	Standard
	Pb	208	47029.6	1.3	**2.3051**	0.014	0.6	ug/L	883	Standard
	U	238	743.4	3.4	**0.0980**	0.002	2.1	ug/L	3	Standard
>	**Bi**	209	552389.3	1.4				ug/L	523984	Standard

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:41:32
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	33.3	17.3	**-0.0260**	0.006	22.5	mg/L	37	Standard
K	39	50.0	36.1	**-0.0133**	0.037	281.6	mg/L	38	Standard
Ca	43	41.7	48.5	**-0.1354**	1.146	846.9	mg/L	10	Standard
Fe	54	2020.3	1.1	**1.4233**	0.051	3.6	mg/L	179	Standard
Fe	57	728.4	10.0	**1.1787**	0.168	14.2	mg/L	335	Standard
> **Sc-1**	45	66977.8	2.5				mg/L	61029	Standard
Cl	35	24720.8	1.1				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	1860.1	4.8				ug/L	1547	Standard
P	31	18349.4	1.0				ug/L	15286	Standard
S	34	1845.1	4.4				ug/L	1572	Standard
Sr	88	65.0	27.7				ug/L	73	Standard
C	12	66.7	31.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	2304.6	5.8				mg/L	6	Standard
Ho-1	165	1405.1	5.7				mg/L	10	Standard
Er	166	1370.1	4.1				mg/L	13	Standard
I	127	4489.0	15.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		107.208	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.514	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:41:32
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.966
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.421
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 15:42:28
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	79858.3	1.9				ug/L	74699	Standard
	Be	9	50198.1	2.5	**53.3484**	1.401	2.6	ug/L	8	Standard
	Al	27	8066966.3	0.3	**51.3427**	0.865	1.7	ug/L	1493	Standard
	Sc	45	68357.2	2.3				ug/L	61029	Standard
	Ti	47	43587.6	0.3	**105.7066**	1.724	1.6	ug/L	35	Standard
	V	51	433103.8	0.8	**52.5653**	0.576	1.1	ug/L	1435	Standard
	Cr	52	431112.3	0.9	**52.4990**	0.789	1.5	ug/L	7635	Standard
	Cr	53	52653.0	1.0	**52.0862**	0.689	1.3	ug/L	363	Standard
	Mn	55	405815.5	1.7	**53.4827**	1.214	2.3	ug/L	747	Standard
	Co	59	390955.7	1.0	**54.4418**	1.354	2.5	ug/L	234	Standard
	Ni	60	105213.1	1.7	**51.5742**	0.488	0.9	ug/L	215	Standard
	Cu	65	104377.6	0.2	**51.4980**	0.905	1.8	ug/L	104	Standard
	Zn	66	61654.9	0.7	**51.4853**	1.289	2.5	ug/L	363	Standard
>	Ge	72	538231.7	1.8				ug/L	495142	Standard
	As	75	61802.6	0.2	**50.8282**	0.881	1.7	ug/L	-50	Standard
	Se	82	6211.6	1.2	**51.9313**	0.364	0.7	ug/L	14	Standard
	Se-1	77	4323.6	2.0	**53.8508**	1.659	3.1	ug/L	42	Standard
>	Ga	71	60.0	28.9				mg/L	42	Standard
	Rb	85	955.0	5.2				ug/L	17	Standard
	Y	89	475770.1	1.3				ug/L	417098	Standard
>	Rh	103	25.0	34.6				ug/L	13	Standard
	Mo	98	290643.5	0.8	**97.6712**	0.653	0.7	ug/L	11	Standard
	Ag	107	354998.2	1.8	**48.7852**	1.288	2.6	ug/L	68	Standard
	Cd	111	115105.3	1.2	**49.3038**	0.151	0.3	mg/L	2	Standard
	Cd	114	278592.4	1.8	**48.6531**	0.360	0.7	ug/L	12	Standard
>	In	115	655825.0	1.0				ug/L	587441	Standard
	Sn	118	319548.2	1.2	**48.9035**	0.256	0.5	ug/L	497	Standard
	Sb	123	249456.0	1.2	**47.8604**	0.919	1.9	ug/L	78	Standard
	Ba	135	108215.4	0.8	**47.3455**	0.546	1.2	ug/L	24	Standard
	Ce	140	256.7	76.3				ug/L	25	Standard
>	Tb	159	1025574.2	1.9				ug/L	939701	Standard
	Ho	165	15.0	0.0				ug/L	10	Standard
	Tl	203	443897.9	1.2	**50.1327**	1.058	2.1	ug/L	4	Standard
	Tl	205	398087.0	0.7	**51.8003**	0.814	1.6	ug/L	2	Standard
	Pb	206	280554.3	0.5	**48.8226**	0.728	1.5	ug/L	240	Standard
	Pb	207	249585.9	0.6	**48.8846**	0.533	1.1	ug/L	198	Standard
	Pb	208	1026713.8	0.6	**50.9050**	1.131	2.2	ug/L	883	Standard
	U	238	372550.1	0.8	**50.3866**	1.127	2.2	ug/L	3	Standard
>	Bi	209	562933.7	1.7				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:44:45
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4292.3	1.6	**4.8821**	0.035	0.7	mg/L	37	Standard
K	39	2345.2	3.1	**4.5133**	0.080	1.8	mg/L	38	Standard
Ca	43	103.3	10.1	**3.3018**	0.528	16.0	mg/L	10	Standard
Fe	54	6681.2	2.2	**4.9206**	0.153	3.1	mg/L	179	Standard
Fe	57	1928.5	4.0	**4.7824**	0.344	7.2	mg/L	335	Standard
Sc-1	45	68357.2	2.3				mg/L	61029	Standard
Cl	35	26753.6	0.7				ug/L	20195	Standard
Kr	83	3.3	45.8				ug/L	4	Standard
Br	81	1833.4	8.1				ug/L	1547	Standard
P	31	20395.4	2.2				ug/L	15286	Standard
S	34	2408.5	8.3				ug/L	1572	Standard
Sr	88	85.0	25.6				ug/L	73	Standard
C	12	53.3	28.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.2	35.5				mg/L	6	Standard
Ho-1	165	15.0	0.0				mg/L	10	Standard
Er	166	30.0	57.7				mg/L	13	Standard
I	127	2536.9	16.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	106.697		
Al	27	102.685		
Sc	45			
Ti	47	105.707		
V	51	105.131		
Cr	52	104.998		
Cr	53			
Mn	55	106.965		
Co	59	108.884		
Ni	60	103.148		
Cu	65	102.996		
Zn	66	102.971		
Ge	72		108.702	
As	75	101.656		
Se	82	103.863		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:44:45
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	97.671		
Ag	107	97.570		
Cd	111	98.608		
Cd	114			
In	115			111.641
Sn	118	97.807		
Sb	123	95.721		
Ba	135	94.691		
Ce	140			
Tb	159			
Ho	165			
Tl	203	100.265		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.810		
U	238	100.773		
Bi	209			107.433
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:44:45
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 15:45:40
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	76618.3	2.3				ug/L	74699	Standard
	Be	9	25.0	40.0	**0.0352**	0.010	29.8	ug/L	8	Standard
	Al	27	2451.9	23.9	**0.0029**	0.004	131.5	ug/L	1493	Standard
	Sc	45	64234.2	1.5				ug/L	61029	Standard
	Ti	47	38.0	9.5	**-0.0165**	0.010	60.4	ug/L	35	Standard
	V	51	1363.5	4.5	**-0.0292**	0.006	19.9	ug/L	1435	Standard
	Cr	52	7102.3	0.4	**-0.1401**	0.016	11.4	ug/L	7635	Standard
	Cr	53	373.3	16.1	**0.0429**	0.063	148.1	ug/L	363	Standard
	Mn	55	1097.0	20.0	**-0.0025**	0.031	1279.7	ug/L	747	Standard
	Co	59	247.3	6.4	**-0.0027**	0.003	101.3	ug/L	234	Standard
	Ni	60	180.3	4.0	**-0.0351**	0.003	7.6	ug/L	215	Standard
	Cu	65	132.0	9.2	**-0.0382**	0.007	18.6	ug/L	104	Standard
	Zn	66	415.3	4.4	**-0.1038**	0.019	18.2	ug/L	363	Standard
>	Ge	72	537099.7	1.7				ug/L	495142	Standard
	As	75	-44.5	23.9	**0.0230**	0.008	35.8	ug/L	-50	Standard
	Se	82	14.3	4.3	**-0.0180**	0.004	23.6	ug/L	14	Standard
	Se-1	77	59.7	7.9	**0.1303**	0.047	35.8	ug/L	42	Standard
>	Ga	71	51.7	29.6				mg/L	42	Standard
	Rb	85	35.0	0.0				ug/L	17	Standard
	Y	89	479758.9	0.4				ug/L	417098	Standard
>	Rh	103	8.3	34.6				ug/L	13	Standard
	Mo	98	277.2	29.2	**0.0885**	0.026	29.3	ug/L	11	Standard
	Ag	107	131.7	17.8	**0.0055**	0.003	58.6	ug/L	68	Standard
	Cd	111	17.2	29.5	**-0.0005**	0.002	465.0	mg/L	2	Standard
	Cd	114	40.7	69.7	**0.0051**	0.005	98.6	ug/L	12	Standard
>	In	115	646048.6	1.6				ug/L	587441	Standard
	Sn	118	843.4	4.5	**0.0366**	0.007	19.0	ug/L	497	Standard
	Sb	123	501.9	29.2	**0.0939**	0.030	32.1	ug/L	78	Standard
	Ba	135	52.0	15.0	**0.0071**	0.003	45.6	ug/L	24	Standard
	Ce	140	70.0	58.5				ug/L	25	Standard
>	Tb	159	979678.9	1.9				ug/L	939701	Standard
	Ho	165	13.3	114.6				ug/L	10	Standard
	Tl	203	68.0	38.6	**0.0062**	0.003	47.0	ug/L	4	Standard
	Tl	205	56.7	27.0	**0.0078**	0.002	25.4	ug/L	2	Standard
	Pb	206	311.7	5.0	**-0.0149**	0.002	16.8	ug/L	240	Standard
	Pb	207	257.0	6.3	**-0.0177**	0.002	12.3	ug/L	198	Standard
	Pb	208	1140.4	10.1	**-0.0154**	0.005	31.9	ug/L	883	Standard
	U	238	55.7	48.1	**0.0033**	0.004	107.8	ug/L	3	Standard
>	Bi	209	542619.7	2.4				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:47:57
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.9504**	31.082	173.2	mg/L	0	Standard
Mg	24	40.0	54.5	**-0.0162**	0.026	161.8	mg/L	37	Standard
K	39	53.3	59.5	**-0.0020**	0.067	3445.3	mg/L	38	Standard
Ca	43	18.3	103.3	**-1.4157**	1.156	81.6	mg/L	10	Standard
Fe	54	130.7	11.8	**-0.0329**	0.014	41.5	mg/L	179	Standard
Fe	57	301.7	9.1	**-0.1005**	0.081	80.3	mg/L	335	Standard
> **Sc-1**	45	64234.2	1.5				mg/L	61029	Standard
Cl	35	26295.5	2.8				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1633.4	7.7				ug/L	1547	Standard
P	31	19669.4	2.1				ug/L	15286	Standard
S	34	2230.2	5.1				ug/L	1572	Standard
Sr	88	70.0	21.4				ug/L	73	Standard
C	12	46.7	61.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	12.2	49.4				mg/L	6	Standard
Ho-1	165	13.3	114.6				mg/L	10	Standard
Er	166	23.3	107.9				mg/L	13	Standard
I	127	1980.1	22.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.474	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:47:57
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.977
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.556
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:47:57
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBW 60 WG574001-04
Sample Date/Time: Wednesday, June 29, 2016 15:48:53
Number of Replicates: 3
Autosampler Position: 205
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79282.0	2.9				ug/L	74699	Standard
	Be	9	31.7	134.3	**0.0407**	0.044	108.3	ug/L	8	Standard
	Al	27	19345.2	59.5	**0.1093**	0.069	63.4	ug/L	1493	Standard
	Sc	45	67500.3	4.8				ug/L	61029	Standard
	Ti	47	51.7	27.5	**0.0181**	0.032	177.8	ug/L	35	Standard
	V	51	1877.4	33.1	**0.0362**	0.072	198.6	ug/L	1435	Standard
	Cr	52	9295.2	4.7	**0.1500**	0.032	21.3	ug/L	7635	Standard
	Cr	53	908.4	6.4	**0.5905**	0.063	10.7	ug/L	363	Standard
	Mn	55	10971.3	111.8	**1.3060**	1.603	122.8	ug/L	747	Standard
	Co	59	555.3	78.4	**0.0407**	0.060	146.5	ug/L	234	Standard
	Ni	60	485.3	47.2	**0.1173**	0.109	92.6	ug/L	215	Standard
	Cu	65	307.3	25.1	**0.0504**	0.035	70.1	ug/L	104	Standard
	Zn	66	1901.5	5.5	**1.1744**	0.056	4.8	ug/L	363	Standard
>	**Ge**	72	528898.5	2.0				ug/L	495142	Standard
	As	75	-0.1	63053.1	**0.0590**	0.043	73.0	ug/L	-50	Standard
	Se	82	20.7	41.6	**0.0373**	0.070	187.2	ug/L	14	Standard
	Se-1	77	69.7	5.4	**0.2713**	0.061	22.4	ug/L	42	Standard
>	**Ga**	71	70.0	75.3				mg/L	42	Standard
	Rb	85	391.7	126.9				ug/L	17	Standard
	Y	89	488721.1	2.3				ug/L	417098	Standard
>	**Rh**	103	16.7	105.4				ug/L	13	Standard
	Mo	98	140.7	71.4	**0.0402**	0.032	80.6	ug/L	11	Standard
	Ag	107	241.3	88.8	**0.0193**	0.028	147.0	ug/L	68	Standard
	Cd	111	97.4	139.8	**0.0324**	0.056	173.4	mg/L	2	Standard
	Cd	114	250.4	134.1	**0.0402**	0.057	140.7	ug/L	12	Standard
>	**In**	115	673056.2	1.0				ug/L	587441	Standard
	Sn	118	918.4	10.7	**0.0424**	0.014	32.3	ug/L	497	Standard
	Sb	123	377.5	33.3	**0.0663**	0.023	34.5	ug/L	78	Standard
	Ba	135	710.7	130.6	**0.2849**	0.392	137.5	ug/L	24	Standard
	Ce	140	2190.4	144.6				ug/L	25	Standard
>	**Tb**	159	1033072.6	2.9				ug/L	939701	Standard
	Ho	165	46.7	118.0				ug/L	10	Standard
	Tl	203	354.0	129.7	**0.0375**	0.050	134.0	ug/L	4	Standard
	Tl	205	175.0	119.3	**0.0226**	0.026	116.3	ug/L	2	Standard
	Pb	206	584.7	65.9	**0.0302**	0.064	211.0	ug/L	240	Standard
	Pb	207	508.0	67.6	**0.0291**	0.064	219.9	ug/L	198	Standard
	Pb	208	1976.7	53.7	**0.0238**	0.049	207.9	ug/L	883	Standard
	U	238	147.7	117.1	**0.0152**	0.023	148.4	ug/L	3	Standard
>	**Bi**	209	558931.4	2.9				ug/L	523984	Standard

Sample ID: PBW 60 WG574001-04
Report Date/Time: Wednesday, June 29, 2016 15:51:10
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	38.3	32.8	**-0.0206**	0.014	67.5	mg/L	37	Standard
K	39	46.7	37.6	**-0.0214**	0.032	149.6	mg/L	38	Standard
Ca	43	20.0	0.0	**-1.3794**	0.055	4.0	mg/L	10	Standard
Fe	54	191.0	21.0	**0.0073**	0.024	325.9	mg/L	179	Standard
Fe	57	261.7	19.2	**-0.2718**	0.132	48.6	mg/L	335	Standard
> **Sc-1**	45	67500.3	4.8				mg/L	61029	Standard
Cl	35	25962.2	2.0				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	2216.8	6.9				ug/L	1547	Standard
P	31	19524.2	2.3				ug/L	15286	Standard
S	34	2248.5	4.1				ug/L	1572	Standard
Sr	88	73.3	27.6				ug/L	73	Standard
C	12	46.7	44.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	67.1	124.3				mg/L	6	Standard
Ho-1	165	46.7	118.0				mg/L	10	Standard
Er	166	60.0	130.2				mg/L	13	Standard
I	127	1978.5	35.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		106.136	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		106.817	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBW 60 WG574001-04
Report Date/Time: Wednesday, June 29, 2016 15:51:10
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.574
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.669
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 60 WG574001-04
Report Date/Time: Wednesday, June 29, 2016 15:51:10
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSW 60 WG574001-05
Sample Date/Time: Wednesday, June 29, 2016 15:52:04
Number of Replicates: 3
Autosampler Position: 206
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	78008.7	2.7				ug/L	74699	Standard
	Be	9	49798.4	3.1	**54.1693**	0.583	1.1	ug/L	8	Standard
	Al	27	8265.6	3.2	**0.0405**	0.001	2.7	ug/L	1493	Standard
	Sc	45	64878.8	4.3				ug/L	61029	Standard
	Ti	47	43.0	16.3	**-0.0042**	0.019	444.7	ug/L	35	Standard
	V	51	424123.5	0.9	**51.5862**	0.653	1.3	ug/L	1435	Standard
	Cr	52	430096.1	0.4	**52.4970**	1.185	2.3	ug/L	7635	Standard
	Cr	53	53307.0	1.3	**52.8599**	1.230	2.3	ug/L	363	Standard
	Mn	55	405428.0	0.3	**53.5518**	1.000	1.9	ug/L	747	Standard
	Co	59	387903.4	0.1	**54.1349**	1.154	2.1	ug/L	234	Standard
	Ni	60	106395.6	1.4	**52.2826**	1.359	2.6	ug/L	215	Standard
	Cu	65	107418.5	1.1	**53.1215**	1.283	2.4	ug/L	104	Standard
	Zn	66	63602.2	0.4	**53.2433**	1.221	2.3	ug/L	363	Standard
>	Ge	72	537059.3	2.0				ug/L	495142	Standard
	As	75	62286.8	1.3	**51.3419**	1.295	2.5	ug/L	-50	Standard
	Se	82	6132.0	1.6	**51.3807**	0.931	1.8	ug/L	14	Standard
	Se-1	77	4215.6	1.2	**52.6015**	0.951	1.8	ug/L	42	Standard
>	Ga	71	76.7	32.8				mg/L	42	Standard
	Rb	85	50.0	36.1				ug/L	17	Standard
	Y	89	473307.8	1.1				ug/L	417098	Standard
>	Rh	103	23.3	75.3				ug/L	13	Standard
	Mo	98	59.9	20.8	**0.0148**	0.004	26.5	ug/L	11	Standard
	Ag	107	358191.0	0.9	**50.3679**	1.706	3.4	ug/L	68	Standard
	Cd	111	112882.0	2.3	**49.4684**	1.332	2.7	mg/L	2	Standard
	Cd	114	273909.8	1.0	**48.9526**	1.484	3.0	ug/L	12	Standard
>	In	115	641215.1	2.7				ug/L	587441	Standard
	Sn	118	785.0	7.1	**0.0286**	0.011	37.9	ug/L	497	Standard
	Sb	123	249794.4	1.0	**49.0383**	1.564	3.2	ug/L	78	Standard
	Ba	135	107708.9	1.4	**48.2176**	1.451	3.0	ug/L	24	Standard
	Ce	140	101.7	27.1				ug/L	25	Standard
>	Tb	159	990038.9	2.7				ug/L	939701	Standard
	Ho	165	11.7	99.0				ug/L	10	Standard
	Tl	203	439976.8	0.4	**51.4932**	0.969	1.9	ug/L	4	Standard
	Tl	205	398605.3	0.6	**53.7491**	0.791	1.5	ug/L	2	Standard
	Pb	206	290681.6	0.4	**52.4241**	0.636	1.2	ug/L	240	Standard
	Pb	207	247225.7	0.4	**50.1841**	0.852	1.7	ug/L	198	Standard
	Pb	208	1030904.4	0.6	**52.9701**	1.178	2.2	ug/L	883	Standard
	U	238	368133.3	0.5	**51.5924**	0.851	1.6	ug/L	3	Standard
>	Bi	209	543222.2	1.6				ug/L	523984	Standard

Sample ID: LCSW 60 WG574001-05
Report Date/Time: Wednesday, June 29, 2016 15:54:21
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	35.0	28.6	**-0.0227**	0.012	52.0	mg/L	37	Standard
K	39	28.3	50.9	**-0.0559**	0.027	48.5	mg/L	38	Standard
Ca	43	35.0	28.6	**-0.4579**	0.506	110.5	mg/L	10	Standard
Fe	54	138.7	22.6	**-0.0277**	0.025	88.8	mg/L	179	Standard
Fe	57	243.3	16.6	**-0.2916**	0.161	55.3	mg/L	335	Standard
> **Sc-1**	45	64878.8	4.3				mg/L	61029	Standard
Cl	35	24949.9	2.3				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	1880.1	6.1				ug/L	1547	Standard
P	31	18362.8	1.6				ug/L	15286	Standard
S	34	2046.8	4.1				ug/L	1572	Standard
Sr	88	70.0	18.9				ug/L	73	Standard
C	12	56.7	56.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	2.2	249.0				mg/L	6	Standard
Ho-1	165	11.7	99.0				mg/L	10	Standard
Er	166	23.3	49.5				mg/L	13	Standard
I	127	1781.8	25.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		104.431	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.466	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 60 WG574001-05
Report Date/Time: Wednesday, June 29, 2016 15:54:21
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.154
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.671
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 60 WG574001-05
Report Date/Time: Wednesday, June 29, 2016 15:54:21
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124702 WG574001-01
Sample Date/Time: Wednesday, June 29, 2016 15:55:16
Number of Replicates: 3
Autosampler Position: 207
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	134635.3	6.0				ug/L	74699	Standard
	Be	9	80.0	18.8	**0.0585**	0.013	21.4	ug/L	8	Standard
	Al	27	38117257.0	1.5	**144.1796**	7.359	5.1	ug/L	1493	Standard
	Sc	45	57575.9	1.8				ug/L	61029	Standard
	Ti	47	164.0	8.6	**0.3117**	0.039	12.5	ug/L	35	Standard
	V	51	2694.7	7.4	**0.1512**	0.027	17.7	ug/L	1435	Standard
	Cr	52	13340.2	1.7	**0.7256**	0.017	2.4	ug/L	7635	Standard
	Cr	53	1556.7	8.2	**1.3078**	0.145	11.1	ug/L	363	Standard
	Mn	55	855520.0	4.5	**119.2592**	4.647	3.9	ug/L	747	Standard
	Co	59	4433.3	4.9	**0.6151**	0.027	4.4	ug/L	234	Standard
	Ni	60	11360.3	1.4	**5.7737**	0.123	2.1	ug/L	215	Standard
	Cu	65	4582.0	1.3	**2.2889**	0.015	0.7	ug/L	104	Standard
	Zn	66	3866.8	1.4	**2.9861**	0.052	1.7	ug/L	363	Standard
>	**Ge**	72	509457.0	0.7				ug/L	495142	Standard
	As	75	332.4	10.8	**0.3479**	0.030	8.6	ug/L	-50	Standard
	Se	82	161.9	7.9	**1.2952**	0.108	8.4	ug/L	14	Standard
	Se-1	77	102.3	14.4	**0.7391**	0.192	26.0	ug/L	42	Standard
>	**Ga**	71	95.0	18.2				mg/L	42	Standard
	Rb	85	141592.2	2.6				ug/L	17	Standard
	Y	89	417616.8	3.6				ug/L	417098	Standard
>	**Rh**	103	198.3	13.9				ug/L	13	Standard
	Mo	98	173.0	3.6	**0.0623**	0.003	5.6	ug/L	11	Standard
	Ag	107	339.3	7.8	**0.0417**	0.005	13.0	ug/L	68	Standard
	Cd	111	139.7	8.2	**0.0622**	0.007	10.5	mg/L	2	Standard
	Cd	114	358.4	9.8	**0.0712**	0.009	12.2	ug/L	12	Standard
>	**In**	115	560534.9	2.1				ug/L	587441	Standard
	Sn	118	620.0	5.6	**0.0166**	0.008	49.6	ug/L	497	Standard
	Sb	123	641.7	15.5	**0.1401**	0.024	17.2	ug/L	78	Standard
	Ba	135	5098.9	2.7	**2.5958**	0.100	3.9	ug/L	24	Standard
	Ce	140	815.0	33.2				ug/L	25	Standard
>	**Tb**	159	928792.3	1.3				ug/L	939701	Standard
	Ho	165	11.7	65.5				ug/L	10	Standard
	Tl	203	1143.7	8.4	**0.1440**	0.014	9.4	ug/L	4	Standard
	Tl	205	1088.4	7.4	**0.1600**	0.012	7.4	ug/L	2	Standard
	Pb	206	578.7	9.2	**0.0425**	0.010	23.1	ug/L	240	Standard
	Pb	207	493.0	9.5	**0.0393**	0.012	31.8	ug/L	198	Standard
	Pb	208	1938.4	3.0	**0.0345**	0.004	12.8	ug/L	883	Standard
	U	238	1219.0	4.3	**0.1816**	0.009	5.1	ug/L	3	Standard
>	**Bi**	209	499083.1	3.3				ug/L	523984	Standard

Sample ID: L1606124702 WG574001-01
Report Date/Time: Wednesday, June 29, 2016 15:57:33
Page 1

Approved: July 04, 2016

Na	23	20.0	43.3	**245.3647**	110.079	44.9	mg/L	0	Standard
Mg	24	80086.1	0.7	**109.5385**	2.239	2.0	mg/L	37	Standard
K	39	1376.7	6.5	**3.1098**	0.163	5.2	mg/L	38	Standard
Ca	43	2828.6	6.5	**186.7611**	8.963	4.8	mg/L	10	Standard
Fe	54	188.7	25.7	**0.0318**	0.047	146.5	mg/L	179	Standard
Fe	57	1751.8	4.9	**5.2348**	0.203	3.9	mg/L	335	Standard
> **Sc-1**	45	57575.9	1.8				mg/L	61029	Standard
Cl	35	65110.3	1.2				ug/L	20195	Standard
Kr	83	3.7	68.6				ug/L	4	Standard
Br	81	22346.5	6.5				ug/L	1547	Standard
P	31	17116.3	3.4				ug/L	15286	Standard
S	34	2223.5	6.9				ug/L	1572	Standard
Sr	88	836.7	13.9				ug/L	73	Standard
C	12	76.7	74.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	52.1	77.4				mg/L	6	Standard
Ho-1	165	11.7	65.5				mg/L	10	Standard
Er	166	26.7	57.3				mg/L	13	Standard
I	127	41219.6	7.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		180.238	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		102.891	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606124702 WG574001-01
Report Date/Time: Wednesday, June 29, 2016 15:57:33
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.420
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.248
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124702 WG574001-01
Report Date/Time: Wednesday, June 29, 2016 15:57:33
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124702DP WG574001-06
Sample Date/Time: Wednesday, June 29, 2016 15:58:27
Number of Replicates: 3
Autosampler Position: 208
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	151702.5	2.0				ug/L	74699	Standard
	Be	9	51.7	59.1	**0.0368**	0.018	48.5	ug/L	8	Standard
	Al	27	41839885.7	0.1	**140.2128**	2.854	2.0	ug/L	1493	Standard
	Sc	45	56811.4	3.1				ug/L	61029	Standard
	Ti	47	131.7	6.1	**0.2545**	0.026	10.2	ug/L	35	Standard
	V	51	2376.3	9.9	**0.1332**	0.026	19.9	ug/L	1435	Standard
	Cr	52	12951.9	1.5	**0.8048**	0.052	6.4	ug/L	7635	Standard
	Cr	53	1565.1	1.7	**1.4417**	0.042	2.9	ug/L	363	Standard
	Mn	55	867683.1	1.0	**130.2493**	3.502	2.7	ug/L	747	Standard
	Co	59	4344.6	3.7	**0.6509**	0.022	3.4	ug/L	234	Standard
	Ni	60	12342.7	0.7	**6.7729**	0.126	1.9	ug/L	215	Standard
	Cu	65	4552.0	1.4	**2.4548**	0.036	1.5	ug/L	104	Standard
	Zn	66	4125.6	2.3	**3.4960**	0.055	1.6	ug/L	363	Standard
>	**Ge**	72	473381.7	2.0				ug/L	495142	Standard
	As	75	389.2	1.9	**0.4231**	0.002	0.6	ug/L	-50	Standard
	Se	82	181.1	4.2	**1.5888**	0.099	6.2	ug/L	14	Standard
	Se-1	77	106.3	18.0	**0.8981**	0.248	27.6	ug/L	42	Standard
>	**Ga**	71	78.3	22.4				mg/L	42	Standard
	Rb	85	147149.5	0.7				ug/L	17	Standard
	Y	89	405581.1	1.8				ug/L	417098	Standard
>	**Rh**	103	178.3	29.5				ug/L	13	Standard
	Mo	98	220.6	12.3	**0.0817**	0.013	15.7	ug/L	11	Standard
	Ag	107	177.0	68.0	**0.0160**	0.020	127.1	ug/L	68	Standard
	Cd	111	65.9	51.9	**0.0256**	0.018	71.0	mg/L	2	Standard
	Cd	114	151.1	49.7	**0.0291**	0.016	55.4	ug/L	12	Standard
>	**In**	115	557347.0	2.4				ug/L	587441	Standard
	Sn	118	766.7	0.4	**0.0437**	0.004	8.6	ug/L	497	Standard
	Sb	123	305.8	17.0	**0.0650**	0.013	20.6	ug/L	78	Standard
	Ba	135	5047.2	3.4	**2.5853**	0.149	5.7	ug/L	24	Standard
	Ce	140	491.7	100.3				ug/L	25	Standard
>	**Tb**	159	935324.5	2.7				ug/L	939701	Standard
	Ho	165	18.3	103.3				ug/L	10	Standard
	Tl	203	947.0	9.8	**0.1179**	0.014	11.5	ug/L	4	Standard
	Tl	205	846.7	27.0	**0.1237**	0.035	28.6	ug/L	2	Standard
	Pb	206	418.3	37.6	**0.0106**	0.032	302.9	ug/L	240	Standard
	Pb	207	351.7	32.0	**0.0074**	0.026	354.3	ug/L	198	Standard
	Pb	208	1404.0	30.0	**0.0041**	0.025	608.6	ug/L	883	Standard
	U	238	972.4	9.3	**0.1427**	0.016	11.1	ug/L	3	Standard
>	**Bi**	209	503603.8	1.5				ug/L	523984	Standard

Sample ID: L1606124702DP WG574001-06
Report Date/Time: Wednesday, June 29, 2016 16:00:44
Page 1

Approved: July 04, 2016

Na	23	25.0	34.6	**311.2656**	112.939	36.3	mg/L	0	Standard
Mg	24	85343.0	0.9	**118.3682**	4.384	3.7	mg/L	37	Standard
K	39	1298.4	2.9	**2.9712**	0.147	4.9	mg/L	38	Standard
Ca	43	2830.3	7.5	**189.4839**	12.043	6.4	mg/L	10	Standard
Fe	54	146.6	14.0	**-0.0044**	0.021	483.2	mg/L	179	Standard
Fe	57	1698.4	7.8	**5.1376**	0.623	12.1	mg/L	335	Standard
> **Sc-1**	45	56811.4	3.1				mg/L	61029	Standard
Cl	35	65999.5	1.4				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	22520.1	6.6				ug/L	1547	Standard
P	31	16799.3	3.7				ug/L	15286	Standard
S	34	2040.1	3.0				ug/L	1572	Standard
Sr	88	920.0	8.9				ug/L	73	Standard
C	12	83.3	25.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	173.2				mg/L	3	Standard
Dy	164	32.1	98.6				mg/L	6	Standard
Ho-1	165	18.3	103.3				mg/L	10	Standard
Er	166	26.7	43.3				mg/L	13	Standard
I	127	45859.6	7.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		203.086	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.605	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606124702DP WG574001-06
Report Date/Time: Wednesday, June 29, 2016 16:00:44
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.877
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.110
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124702DP WG574001-06
Report Date/Time: Wednesday, June 29, 2016 16:00:44
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126501 WG574001-02
Sample Date/Time: Wednesday, June 29, 2016 16:01:39
Number of Replicates: 3
Autosampler Position: 209
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	112017.9	3.5				ug/L	74699	Standard
	Be	9	21.7	35.3	**0.0242**	0.006	26.3	ug/L	8	Standard
	Al	27	58851331.1	1.0	**267.2037**	7.701	2.9	ug/L	1493	Standard
	Sc	45	63265.2	2.1				ug/L	61029	Standard
	Ti	47	314.3	6.0	**0.7250**	0.050	6.9	ug/L	35	Standard
	V	51	84379.6	0.6	**11.0401**	0.098	0.9	ug/L	1435	Standard
	Cr	52	25344.2	1.7	**2.4155**	0.065	2.7	ug/L	7635	Standard
	Cr	53	30644.3	3.7	**33.0161**	1.295	3.9	ug/L	363	Standard
	Mn	55	3058.3	0.3	**0.2937**	0.003	0.9	ug/L	747	Standard
	Co	59	981.4	1.2	**0.1122**	0.002	1.4	ug/L	234	Standard
	Ni	60	4598.7	1.8	**2.3463**	0.051	2.2	ug/L	215	Standard
	Cu	65	1226.7	2.1	**0.5592**	0.015	2.8	ug/L	104	Standard
	Zn	66	2978.0	3.2	**2.2873**	0.091	4.0	ug/L	363	Standard
>	**Ge**	72	492356.2	0.3				ug/L	495142	Standard
	As	75	3467.1	5.3	**3.1728**	0.175	5.5	ug/L	-50	Standard
	Se	82	1561.8	3.2	**14.1732**	0.483	3.4	ug/L	14	Standard
	Se-1	77	3085.3	7.4	**41.8633**	3.228	7.7	ug/L	42	Standard
>	**Ga**	71	63.3	27.7				mg/L	42	Standard
	Rb	85	9206.1	3.0				ug/L	17	Standard
	Y	89	442169.1	3.5				ug/L	417098	Standard
>	**Rh**	103	255.0	11.9				ug/L	13	Standard
	Mo	98	5205.2	3.3	**1.8993**	0.033	1.7	ug/L	11	Standard
	Ag	107	803.7	5.3	**0.1074**	0.005	4.8	ug/L	68	Standard
	Cd	111	45.0	9.2	**0.0131**	0.002	16.6	mg/L	2	Standard
	Cd	114	163.3	22.8	**0.0289**	0.007	22.9	ug/L	12	Standard
>	**In**	115	602229.8	3.0				ug/L	587441	Standard
	Sn	118	683.3	9.8	**0.0193**	0.009	45.4	ug/L	497	Standard
	Sb	123	210.7	19.8	**0.0397**	0.008	19.9	ug/L	78	Standard
	Ba	135	21407.4	0.7	**10.1914**	0.253	2.5	ug/L	24	Standard
	Ce	140	328.3	10.1				ug/L	25	Standard
>	**Tb**	159	984916.0	0.2				ug/L	939701	Standard
	Ho	165	35.0	24.7				ug/L	10	Standard
	Tl	203	94.3	9.0	**0.0098**	0.001	12.0	ug/L	4	Standard
	Tl	205	65.0	40.7	**0.0094**	0.004	40.6	ug/L	2	Standard
	Pb	206	401.3	3.0	**0.0049**	0.001	28.2	ug/L	240	Standard
	Pb	207	354.0	5.2	**0.0056**	0.003	55.2	ug/L	198	Standard
	Pb	208	1436.0	3.5	**0.0035**	0.003	74.9	ug/L	883	Standard
	U	238	31993.4	1.5	**4.7067**	0.092	1.9	ug/L	3	Standard
>	**Bi**	209	517006.9	1.2				ug/L	523984	Standard

Sample ID: L1606126501 WG574001-02
Report Date/Time: Wednesday, June 29, 2016 16:03:56
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**55.9813**	56.756	101.4	mg/L	0	Standard
Mg	24	180057.1	1.6	**224.1644**	1.858	0.8	mg/L	37	Standard
K	39	641.7	5.8	**1.2549**	0.096	7.6	mg/L	38	Standard
Ca	43	708.3	7.0	**40.6474**	3.060	7.5	mg/L	10	Standard
Fe	54	157.9	5.8	**-0.0091**	0.008	86.4	mg/L	179	Standard
Fe	57	550.0	17.3	**0.7314**	0.333	45.5	mg/L	335	Standard
> **Sc-1**	45	63265.2	2.1				mg/L	61029	Standard
Cl	35	33676.1	1.7				ug/L	20195	Standard
Kr	83	0.3	173.2				ug/L	4	Standard
Br	81	48808.6	5.5				ug/L	1547	Standard
P	31	23084.3	3.5				ug/L	15286	Standard
S	34	2106.8	4.3				ug/L	1572	Standard
Sr	88	1195.0	6.7				ug/L	73	Standard
C	12	130.0	40.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	9.0	113.3				mg/L	6	Standard
Ho-1	165	35.0	24.7				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	49504.2	3.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		149.959	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.437	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126501 WG574001-02
Report Date/Time: Wednesday, June 29, 2016 16:03:56
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.517
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.668
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606126501 WG574001-02
Report Date/Time: Wednesday, June 29, 2016 16:03:56
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126502S WG574001-07
Sample Date/Time: Wednesday, June 29, 2016 16:04:50
Number of Replicates: 3
Autosampler Position: 210
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	109865.8	4.5				ug/L	74699	Standard
	Be	9	61395.8	1.0	**47.4912**	2.348	4.9	ug/L	8	Standard
	Al	27	61297739.1	1.7	**284.0961**	17.361	6.1	ug/L	1493	Standard
	Sc	45	64222.6	2.9				ug/L	61029	Standard
	Ti	47	334.7	6.1	**0.7462**	0.065	8.7	ug/L	35	Standard
	V	51	514269.3	0.9	**65.7128**	0.488	0.7	ug/L	1435	Standard
	Cr	52	414721.1	1.2	**53.1400**	0.316	0.6	ug/L	7635	Standard
	Cr	53	79122.8	2.2	**82.5246**	0.802	1.0	ug/L	363	Standard
	Mn	55	379075.2	0.4	**52.5566**	0.801	1.5	ug/L	747	Standard
	Co	59	375122.1	1.7	**54.9533**	1.379	2.5	ug/L	234	Standard
	Ni	60	98886.4	1.8	**50.9941**	0.435	0.9	ug/L	215	Standard
	Cu	65	94781.9	0.9	**49.1885**	0.428	0.9	ug/L	104	Standard
	Zn	66	59708.1	1.0	**52.4531**	0.349	0.7	ug/L	363	Standard
>	Ge	72	511580.8	1.5				ug/L	495142	Standard
	As	75	64226.1	0.6	**55.5617**	0.582	1.0	ug/L	-50	Standard
	Se	82	7637.7	1.0	**67.2251**	1.117	1.7	ug/L	14	Standard
	Se-1	77	7177.4	2.9	**94.4764**	1.329	1.4	ug/L	42	Standard
>	Ga	71	96.7	7.9				mg/L	42	Standard
	Rb	85	9571.4	0.6				ug/L	17	Standard
	Y	89	459795.2	1.7				ug/L	417098	Standard
>	Rh	103	293.3	12.0				ug/L	13	Standard
	Mo	98	5421.1	3.7	**1.9704**	0.066	3.4	ug/L	11	Standard
	Ag	107	318338.0	1.0	**47.4485**	0.356	0.8	ug/L	68	Standard
	Cd	111	104991.2	1.0	**48.7852**	0.464	1.0	mg/L	2	Standard
	Cd	114	250399.7	1.5	**47.4405**	0.760	1.6	ug/L	12	Standard
>	In	115	604563.1	0.3				ug/L	587441	Standard
	Sn	118	798.4	4.4	**0.0381**	0.006	16.0	ug/L	497	Standard
	Sb	123	240747.3	1.0	**50.1002**	0.397	0.8	ug/L	78	Standard
	Ba	135	122571.5	0.3	**58.1746**	0.379	0.7	ug/L	24	Standard
	Ce	140	328.3	0.9				ug/L	25	Standard
>	Tb	159	981221.4	1.4				ug/L	939701	Standard
	Ho	165	28.3	44.4				ug/L	10	Standard
	Tl	203	414235.7	0.7	**51.0974**	0.645	1.3	ug/L	4	Standard
	Tl	205	378219.2	0.6	**53.7534**	0.090	0.2	ug/L	2	Standard
	Pb	206	277784.6	0.1	**52.8053**	0.305	0.6	ug/L	240	Standard
	Pb	207	237018.9	0.5	**50.7105**	0.514	1.0	ug/L	198	Standard
	Pb	208	965977.0	0.4	**52.3096**	0.506	1.0	ug/L	883	Standard
	U	238	407586.9	0.7	**60.2072**	0.700	1.2	ug/L	3	Standard
>	Bi	209	515321.0	0.6				ug/L	523984	Standard

Sample ID: L1606126502S WG574001-07
Report Date/Time: Wednesday, June 29, 2016 16:07:07
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.6720**	30.600	173.2	mg/L	0	Standard
Mg	24	189285.1	1.3	**232.3057**	9.548	4.1	mg/L	37	Standard
K	39	641.7	6.1	**1.2328**	0.045	3.6	mg/L	38	Standard
Ca	43	760.0	12.3	**43.0240**	4.391	10.2	mg/L	10	Standard
Fe	54	185.4	16.8	**0.0114**	0.028	245.2	mg/L	179	Standard
Fe	57	630.0	4.2	**0.9612**	0.087	9.0	mg/L	335	Standard
> **Sc-1**	45	64222.6	2.9				mg/L	61029	Standard
Cl	35	34933.6	1.5				ug/L	20195	Standard
Kr	83	2.7	114.6				ug/L	4	Standard
Br	81	53306.9	1.0				ug/L	1547	Standard
P	31	23740.3	1.4				ug/L	15286	Standard
S	34	2118.5	3.3				ug/L	1572	Standard
Sr	88	1145.0	3.3				ug/L	73	Standard
C	12	140.0	21.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	28.6	38.4				mg/L	6	Standard
Ho-1	165	28.3	44.4				mg/L	10	Standard
Er	166	30.0	66.7				mg/L	13	Standard
I	127	81496.0	7.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		147.078	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.320	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126502S WG574001-07
Report Date/Time: Wednesday, June 29, 2016 16:07:07
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	102.915	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	98.347	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606126502S WG574001-07
Report Date/Time: Wednesday, June 29, 2016 16:07:07
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126503SD WG574001-08
Sample Date/Time: Wednesday, June 29, 2016 16:08:02
Number of Replicates: 3
Autosampler Position: 211
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	108017.3	3.5				ug/L	74699	Standard
	Be	9	61104.9	5.2	**47.9870**	0.846	1.8	ug/L	8	Standard
	Al	27	60436124.3	2.1	**284.5445**	9.073	3.2	ug/L	1493	Standard
	Sc	45	65195.1	3.6				ug/L	61029	Standard
	Ti	47	321.3	5.9	**0.6738**	0.054	8.0	ug/L	35	Standard
	V	51	521782.6	1.8	**63.5667**	1.206	1.9	ug/L	1435	Standard
	Cr	52	431117.5	0.3	**52.6702**	0.983	1.9	ug/L	7635	Standard
	Cr	53	81998.0	1.1	**81.5509**	0.729	0.9	ug/L	363	Standard
	Mn	55	379421.8	1.7	**50.1647**	1.838	3.7	ug/L	747	Standard
	Co	59	378746.6	2.0	**52.9167**	2.098	4.0	ug/L	234	Standard
	Ni	60	101643.5	0.7	**49.9872**	1.144	2.3	ug/L	215	Standard
	Cu	65	96036.4	1.2	**47.5169**	0.455	1.0	ug/L	104	Standard
	Zn	66	60629.4	1.5	**50.7689**	0.479	0.9	ug/L	363	Standard
>	**Ge**	72	536566.5	1.9				ug/L	495142	Standard
	As	75	65668.7	0.7	**54.1701**	0.796	1.5	ug/L	-50	Standard
	Se	82	7769.6	1.1	**65.2136**	1.970	3.0	ug/L	14	Standard
	Se-1	77	7316.4	1.6	**91.8340**	1.824	2.0	ug/L	42	Standard
>	**Ga**	71	56.7	22.2				mg/L	42	Standard
	Rb	85	9684.8	3.9				ug/L	17	Standard
	Y	89	458590.9	3.1				ug/L	417098	Standard
>	**Rh**	103	275.0	12.6				ug/L	13	Standard
	Mo	98	5397.9	1.9	**1.9792**	0.040	2.0	ug/L	11	Standard
	Ag	107	315299.7	0.9	**47.4082**	1.008	2.1	ug/L	68	Standard
	Cd	111	105649.0	0.5	**49.5183**	0.751	1.5	mg/L	2	Standard
	Cd	114	256574.5	1.5	**49.0334**	1.028	2.1	ug/L	12	Standard
>	**In**	115	599465.1	2.0				ug/L	587441	Standard
	Sn	118	820.0	11.4	**0.0428**	0.015	34.4	ug/L	497	Standard
	Sb	123	245153.5	0.2	**51.4644**	0.950	1.8	ug/L	78	Standard
	Ba	135	122768.4	0.8	**58.7829**	1.524	2.6	ug/L	24	Standard
	Ce	140	333.3	10.5				ug/L	25	Standard
>	**Tb**	159	983758.7	0.8				ug/L	939701	Standard
	Ho	165	18.3	41.7				ug/L	10	Standard
	Tl	203	416470.0	0.7	**50.8907**	0.470	0.9	ug/L	4	Standard
	Tl	205	377588.7	0.1	**53.1619**	0.172	0.3	ug/L	2	Standard
	Pb	206	279930.3	0.5	**52.7144**	0.351	0.7	ug/L	240	Standard
	Pb	207	238634.3	0.7	**50.5762**	0.275	0.5	ug/L	198	Standard
	Pb	208	958199.8	1.0	**51.4011**	0.688	1.3	ug/L	883	Standard
	U	238	405052.0	0.8	**59.2716**	0.671	1.1	ug/L	3	Standard
>	**Bi**	209	520189.0	0.3				ug/L	523984	Standard

Sample ID: L1606126503SD WG574001-08
Report Date/Time: Wednesday, June 29, 2016 16:10:18
Page 1

Approved: July 04, 2016

Na	23	5.0	173.2	**51.9909**	90.042	173.2	mg/L	0	Standard
Mg	24	184795.4	2.1	**223.5320**	12.550	5.6	mg/L	37	Standard
K	39	668.3	13.7	**1.2668**	0.164	13.0	mg/L	38	Standard
Ca	43	725.0	7.3	**40.3567**	2.975	7.4	mg/L	10	Standard
Fe	54	151.4	21.3	**-0.0183**	0.024	132.3	mg/L	179	Standard
Fe	57	623.3	16.5	**0.9045**	0.268	29.6	mg/L	335	Standard
> **Sc-1**	45	65195.1	3.6				mg/L	61029	Standard
Cl	35	33686.7	1.5				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	51370.3	3.2				ug/L	1547	Standard
P	31	23124.3	1.4				ug/L	15286	Standard
S	34	2101.8	2.3				ug/L	1572	Standard
Sr	88	1251.7	9.6				ug/L	73	Standard
C	12	163.3	19.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	9.4	104.3				mg/L	6	Standard
Ho-1	165	18.3	41.7				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	90519.1	8.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		144.604	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.366	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126503SD WG574001-08
Report Date/Time: Wednesday, June 29, 2016 16:10:18
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	102.047	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	99.276	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701
Sample Date/Time: Wednesday, June 29, 2016 16:11:13
Number of Replicates: 3
Autosampler Position: 212
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	160464.4	3.8				ug/L	74699	Standard
	Be	9	12485.1	1.8	**6.6126**	0.132	2.0	ug/L	8	Standard
	Al	27	129371635.5	1.7	**410.0067**	8.683	2.1	ug/L	1493	Standard
	Sc	45	97873.5	2.0				ug/L	61029	Standard
	Ti	47	12940.2	2.9	**32.4229**	1.439	4.4	ug/L	35	Standard
	V	51	62587.9	0.8	**7.6988**	0.173	2.3	ug/L	1435	Standard
	Cr	52	130472.4	1.2	**15.7449**	0.176	1.1	ug/L	7635	Standard
	Cr	53	17299.9	2.8	**17.4949**	0.102	0.6	ug/L	363	Standard
	Mn	55	53885293.3	1.0	**7372.4056**	142.217	1.9	ug/L	747	Standard
	Co	59	1486809.2	2.0	**214.4609**	6.754	3.1	ug/L	234	Standard
	Ni	60	953927.1	1.9	**485.1518**	7.673	1.6	ug/L	215	Standard
	Cu	65	14021.8	2.7	**7.0705**	0.048	0.7	ug/L	104	Standard
	Zn	66	1250022.8	1.4	**1089.5042**	11.046	1.0	ug/L	363	Standard
>	**Ge**	72	519921.5	2.2				ug/L	495142	Standard
	As	75	2369.9	1.5	**2.0748**	0.031	1.5	ug/L	-50	Standard
	Se	82	334.2	2.8	**2.7639**	0.140	5.1	ug/L	14	Standard
	Se-1	77	477.7	9.4	**5.6016**	0.492	8.8	ug/L	42	Standard
>	**Ga**	71	3503.7	1.4				mg/L	42	Standard
	Rb	85	155749.9	1.6				ug/L	17	Standard
	Y	89	3051185.4	0.7				ug/L	417098	Standard
>	**Rh**	103	200.0	9.0				ug/L	13	Standard
	Mo	98	648.1	0.8	**0.2277**	0.002	1.0	ug/L	11	Standard
	Ag	107	406.3	10.5	**0.0469**	0.006	13.8	ug/L	68	Standard
	Cd	111	4037.2	2.2	**1.8458**	0.056	3.0	mg/L	2	Standard
	Cd	114	10321.7	4.5	**1.9292**	0.065	3.4	ug/L	12	Standard
>	**In**	115	612067.4	1.8				ug/L	587441	Standard
	Sn	118	936.7	6.3	**0.0593**	0.012	20.5	ug/L	497	Standard
	Sb	123	674.3	5.8	**0.1344**	0.006	4.1	ug/L	78	Standard
	Ba	135	15717.8	1.3	**7.3556**	0.136	1.8	ug/L	24	Standard
	Ce	140	2639736.8	0.7				ug/L	25	Standard
>	**Tb**	159	1213429.7	1.8				ug/L	939701	Standard
	Ho	165	212831.1	0.7				ug/L	10	Standard
	Tl	203	2423.5	3.4	**0.2853**	0.011	3.9	ug/L	4	Standard
	Tl	205	2225.2	6.3	**0.3038**	0.017	5.8	ug/L	2	Standard
	Pb	206	7086.3	0.1	**1.2240**	0.010	0.8	ug/L	240	Standard
	Pb	207	5547.0	2.2	**1.0710**	0.022	2.1	ug/L	198	Standard
	Pb	208	23218.6	1.0	**1.1348**	0.005	0.5	ug/L	883	Standard
	U	238	25740.2	1.8	**3.6463**	0.044	1.2	ug/L	3	Standard
>	**Bi**	209	536711.9	0.7				ug/L	523984	Standard

Sample ID: L1606124701
Report Date/Time: Wednesday, June 29, 2016 16:13:30
Page 1

Approved: July 04, 2016

Na	23	40.0	57.3	**285.3574**	160.270	56.2	mg/L	0	Standard
Mg	24	87773.6	1.9	**70.5914**	1.090	1.5	mg/L	37	Standard
K	39	1211.7	2.8	**1.5562**	0.053	3.4	mg/L	38	Standard
Ca	43	3308.7	4.2	**127.7939**	4.640	3.6	mg/L	10	Standard
Fe	54	180981.2	2.7	**95.5321**	0.892	0.9	mg/L	179	Standard
Fe	57	48873.8	3.4	**102.4969**	2.098	2.0	mg/L	335	Standard
Sc-1	45	97873.5	2.0				mg/L	61029	Standard
Cl	35	64420.7	1.3				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	26873.8	3.7				ug/L	1547	Standard
P	31	52208.1	1.4				ug/L	15286	Standard
S	34	1956.8	2.0				ug/L	1572	Standard
Sr	88	1171.7	10.6				ug/L	73	Standard
C	12	126.7	46.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	329920.1	1.3				mg/L	6	Standard
Ho-1	165	212831.1	0.7				mg/L	10	Standard
Er	166	185158.1	1.8				mg/L	13	Standard
I	127	88143.1	6.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		214.815	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		105.004	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606124701
Report Date/Time: Wednesday, June 29, 2016 16:13:30
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.192
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.429
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701
Report Date/Time: Wednesday, June 29, 2016 16:13:30
Page 3

Approved: July 04, 2016

Co 59 Upper, S, EEE	Co	59
Ni 60 Upper, S, EEE	Ni	60
Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701PS WG574688-01
Sample Date/Time: Wednesday, June 29, 2016 16:14:25
Number of Replicates: 3
Autosampler Position: 213
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	166369.0	3.1				ug/L	74699	Standard
	Be	9	78228.1	1.5	**39.9183**	0.744	1.9	ug/L	8	Standard
	Al	27	127858501.0	1.9	**390.7420**	6.863	1.8	ug/L	1493	Standard
	Sc	45	95075.8	2.0				ug/L	61029	Standard
	Ti	47	12076.5	3.4	**30.6847**	0.840	2.7	ug/L	35	Standard
	V	51	487373.4	0.7	**62.1823**	0.880	1.4	ug/L	1435	Standard
	Cr	52	506749.0	0.5	**65.0704**	0.602	0.9	ug/L	7635	Standard
	Cr	53	63546.3	1.7	**66.1298**	1.201	1.8	ug/L	363	Standard
	Mn	55	53730576.4	0.3	**7459.5309**	74.503	1.0	ug/L	747	Standard
	Co	59	1866113.9	0.2	**273.1251**	3.396	1.2	ug/L	234	Standard
	Ni	60	1009610.1	0.8	**521.0961**	4.502	0.9	ug/L	215	Standard
	Cu	65	102190.2	1.2	**52.9646**	0.037	0.1	ug/L	104	Standard
	Zn	66	1265218.7	0.3	**1119.1475**	11.121	1.0	ug/L	363	Standard
>	**Ge**	72	512284.2	1.2				ug/L	495142	Standard
	As	75	65072.9	0.3	**56.2148**	0.511	0.9	ug/L	-50	Standard
	Se	82	7179.2	1.0	**63.0878**	0.484	0.8	ug/L	14	Standard
	Se-1	77	4836.8	0.7	**63.3901**	0.749	1.2	ug/L	42	Standard
>	**Ga**	71	3398.7	7.2				mg/L	42	Standard
	Rb	85	158551.9	0.5				ug/L	17	Standard
	Y	89	3079691.2	1.3				ug/L	417098	Standard
>	**Rh**	103	243.3	6.3				ug/L	13	Standard
	Mo	98	610.5	4.6	**0.2160**	0.008	3.7	ug/L	11	Standard
	Ag	107	329069.5	0.6	**48.8691**	0.577	1.2	ug/L	68	Standard
	Cd	111	109320.3	0.2	**50.6101**	0.459	0.9	mg/L	2	Standard
	Cd	114	265491.9	0.4	**50.1152**	0.608	1.2	ug/L	12	Standard
>	**In**	115	606831.5	1.0				ug/L	587441	Standard
	Sn	118	1098.4	5.9	**0.0874**	0.012	14.0	ug/L	497	Standard
	Sb	123	240923.7	1.8	**49.9558**	1.171	2.3	ug/L	78	Standard
	Ba	135	113288.6	0.8	**53.5705**	0.782	1.5	ug/L	24	Standard
	Ce	140	2588807.2	1.1				ug/L	25	Standard
>	**Tb**	159	1203377.0	1.6				ug/L	939701	Standard
	Ho	165	210409.6	1.8				ug/L	10	Standard
	Tl	203	437720.2	1.4	**52.3305**	0.744	1.4	ug/L	4	Standard
	Tl	205	404409.4	0.6	**55.7069**	0.222	0.4	ug/L	2	Standard
	Pb	206	295264.9	0.3	**54.4023**	0.212	0.4	ug/L	240	Standard
	Pb	207	250626.5	0.1	**51.9726**	0.335	0.6	ug/L	198	Standard
	Pb	208	1033784.3	0.4	**54.2614**	0.547	1.0	ug/L	883	Standard
	U	238	423217.8	0.6	**60.5919**	0.666	1.1	ug/L	3	Standard
>	**Bi**	209	531687.2	0.6				ug/L	523984	Standard

Sample ID: L1606124701PS WG574688-01
Report Date/Time: Wednesday, June 29, 2016 16:16:41
Page 1

Approved: July 04, 2016

Na	23	31.7	32.9	**235.3667**	80.946	34.4	mg/L	0	Standard
Mg	24	88559.7	0.9	**73.3333**	1.498	2.0	mg/L	37	Standard
K	39	1311.7	6.5	**1.7460**	0.092	5.3	mg/L	38	Standard
Ca	43	3437.1	1.4	**136.8797**	3.742	2.7	mg/L	10	Standard
Fe	54	183063.9	0.7	**99.5077**	1.370	1.4	mg/L	179	Standard
Fe	57	49900.4	2.7	**107.8438**	4.358	4.0	mg/L	335	Standard
> **Sc-1**	45	95075.8	2.0				mg/L	61029	Standard
Cl	35	63889.8	1.2				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	26186.0	7.9				ug/L	1547	Standard
P	31	52633.0	2.3				ug/L	15286	Standard
S	34	1903.5	2.7				ug/L	1572	Standard
Sr	88	1258.4	8.1				ug/L	73	Standard
C	12	153.3	16.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	329580.7	1.9				mg/L	6	Standard
Ho-1	165	210409.6	1.8				mg/L	10	Standard
Er	166	182077.6	1.3				mg/L	13	Standard
I	127	87547.6	5.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		222.720	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.462	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606124701PS WG574688-01
Report Date/Time: Wednesday, June 29, 2016 16:16:41
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	103.301
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	101.470
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701PS WG574688-01
Report Date/Time: Wednesday, June 29, 2016 16:16:41
Page 3

Approved: July 04, 2016

Co 59 Upper, S, EEE	Co	59
Ni 60 Upper, S, EEE	Ni	60
Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701SDL WG574688-02
Sample Date/Time: Wednesday, June 29, 2016 16:17:36
Number of Replicates: 3
Autosampler Position: 214
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	115043.6	2.7				ug/L	74699	Standard
	Be	9	2510.2	6.0	**1.8579**	0.064	3.4	ug/L	8	Standard
	Al	27	26603666.2	1.3	**117.5773**	3.118	2.7	ug/L	1493	Standard
	Sc	45	70520.4	2.0				ug/L	61029	Standard
	Ti	47	2655.2	2.4	**6.2496**	0.159	2.5	ug/L	35	Standard
	V	51	13651.6	1.3	**1.4451**	0.028	2.0	ug/L	1435	Standard
	Cr	52	32029.5	0.6	**2.9000**	0.113	3.9	ug/L	7635	Standard
	Cr	53	4062.2	6.0	**3.6565**	0.147	4.0	ug/L	363	Standard
	Mn	55	11326454.5	1.3	**1476.9710**	52.700	3.6	ug/L	747	Standard
	Co	59	315829.2	0.8	**43.3845**	1.293	3.0	ug/L	234	Standard
	Ni	60	206220.6	0.5	**99.8677**	2.554	2.6	ug/L	215	Standard
	Cu	65	3227.7	3.3	**1.4714**	0.087	5.9	ug/L	104	Standard
	Zn	66	276031.8	0.8	**228.9686**	6.765	3.0	ug/L	363	Standard
>	**Ge**	72	545622.5	2.3				ug/L	495142	Standard
	As	75	529.0	5.1	**0.4886**	0.031	6.3	ug/L	-50	Standard
	Se	82	109.8	7.8	**0.7692**	0.061	7.9	ug/L	14	Standard
	Se-1	77	127.7	10.6	**0.9669**	0.203	21.0	ug/L	42	Standard
>	**Ga**	71	896.7	4.3				mg/L	42	Standard
	Rb	85	33246.4	0.8				ug/L	17	Standard
	Y	89	1035711.2	2.7				ug/L	417098	Standard
>	**Rh**	103	68.3	33.0				ug/L	13	Standard
	Mo	98	137.2	4.4	**0.0416**	0.001	3.0	ug/L	11	Standard
	Ag	107	193.7	13.0	**0.0144**	0.003	20.5	ug/L	68	Standard
	Cd	111	922.1	1.5	**0.3978**	0.011	2.7	mg/L	2	Standard
	Cd	114	2202.4	8.0	**0.3929**	0.032	8.2	ug/L	12	Standard
>	**In**	115	638835.9	2.4				ug/L	587441	Standard
	Sn	118	601.7	4.7	**-0.0000**	0.004	76937.0	ug/L	497	Standard
	Sb	123	1960.6	22.3	**0.3834**	0.095	24.7	ug/L	78	Standard
	Ba	135	3410.7	1.0	**1.5171**	0.049	3.2	ug/L	24	Standard
	Ce	140	553030.9	1.0				ug/L	25	Standard
>	**Tb**	159	1064988.1	1.7				ug/L	939701	Standard
	Ho	165	44299.7	1.5				ug/L	10	Standard
	Tl	203	596.3	8.1	**0.0662**	0.006	8.4	ug/L	4	Standard
	Tl	205	613.3	10.1	**0.0808**	0.008	10.1	ug/L	2	Standard
	Pb	206	1759.1	0.5	**0.2383**	0.002	0.8	ug/L	240	Standard
	Pb	207	1407.7	5.3	**0.2088**	0.015	7.2	ug/L	198	Standard
	Pb	208	5898.4	2.6	**0.2216**	0.008	3.7	ug/L	883	Standard
	U	238	5357.3	1.2	**0.7271**	0.010	1.3	ug/L	3	Standard
>	**Bi**	209	557502.4	0.2				ug/L	523984	Standard

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:19:53
Page 1

Approved: July 04, 2016

Na	23	11.7	89.2	**117.5100**	105.501	89.8	mg/L	0	Standard
Mg	24	18105.8	1.6	**20.1624**	0.144	0.7	mg/L	37	Standard
K	39	321.7	3.6	**0.5016**	0.035	6.9	mg/L	38	Standard
Ca	43	733.4	8.9	**37.5359**	2.964	7.9	mg/L	10	Standard
Fe	54	38163.2	2.6	**27.8607**	0.186	0.7	mg/L	179	Standard
Fe	57	10221.8	1.3	**28.9959**	0.316	1.1	mg/L	335	Standard
> **Sc-1**	45	70520.4	2.0				mg/L	61029	Standard
Cl	35	31180.1	1.1				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	7325.1	3.6				ug/L	1547	Standard
P	31	24873.7	0.5				ug/L	15286	Standard
S	34	1646.8	8.5				ug/L	1572	Standard
Sr	88	296.7	11.0				ug/L	73	Standard
C	12	96.7	53.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	69181.7	2.0				mg/L	6	Standard
Ho-1	165	44299.7	1.5				mg/L	10	Standard
Er	166	39063.6	1.3				mg/L	13	Standard
I	127	21375.1	3.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		154.010	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.195	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:19:53
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.749
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.397
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:19:53
Page 3

Approved: July 04, 2016

Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 16:20:49
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	104492.8	3.9				ug/L	74699	Standard
	Be	9	63869.3	1.0	**51.9084**	1.516	2.9	ug/L	8	Standard
	Al	27	9148865.9	2.6	**44.5067**	0.724	1.6	ug/L	1493	Standard
	Sc	45	68447.7	3.2				ug/L	61029	Standard
	Ti	47	43923.6	1.6	**106.2047**	1.334	1.3	ug/L	35	Standard
	V	51	454775.2	0.5	**55.0469**	0.468	0.8	ug/L	1435	Standard
	Cr	52	431898.3	0.6	**52.4410**	0.316	0.6	ug/L	7635	Standard
	Cr	53	53343.8	2.8	**52.6203**	1.459	2.8	ug/L	363	Standard
	Mn	55	427711.0	4.2	**56.2006**	1.985	3.5	ug/L	747	Standard
	Co	59	425402.4	1.1	**59.0619**	0.614	1.0	ug/L	234	Standard
	Ni	60	106741.4	0.9	**52.1773**	0.185	0.4	ug/L	215	Standard
	Cu	65	106457.4	0.4	**52.3703**	0.176	0.3	ug/L	104	Standard
	Zn	66	63607.9	0.7	**52.9682**	0.530	1.0	ug/L	363	Standard
>	Ge	72	539725.1	0.7				ug/L	495142	Standard
	As	75	63945.9	1.5	**52.4378**	1.123	2.1	ug/L	-50	Standard
	Se	82	6584.0	0.6	**54.8988**	0.684	1.2	ug/L	14	Standard
	Se-1	77	4386.3	1.4	**54.4772**	1.111	2.0	ug/L	42	Standard
>	Ga	71	80.0	16.5				mg/L	42	Standard
	Rb	85	970.0	8.8				ug/L	17	Standard
	Y	89	494238.2	2.5				ug/L	417098	Standard
>	Rh	103	45.0	11.1				ug/L	13	Standard
	Mo	98	317512.9	0.8	**102.4795**	1.834	1.8	ug/L	11	Standard
	Ag	107	378789.7	0.9	**49.9860**	0.591	1.2	ug/L	68	Standard
	Cd	111	119686.6	0.4	**49.2403**	0.865	1.8	mg/L	2	Standard
	Cd	114	291890.5	1.4	**48.9669**	1.322	2.7	ug/L	12	Standard
>	In	115	682927.0	1.5				ug/L	587441	Standard
	Sn	118	338721.1	0.8	**49.7884**	0.723	1.5	ug/L	497	Standard
	Sb	123	263759.6	1.2	**48.5931**	0.331	0.7	ug/L	78	Standard
	Ba	135	111614.3	0.7	**46.9010**	0.972	2.1	ug/L	24	Standard
	Ce	140	1265.1	98.5				ug/L	25	Standard
>	Tb	159	1064058.1	1.0				ug/L	939701	Standard
	Ho	165	88.3	85.2				ug/L	10	Standard
	Tl	203	475263.7	0.7	**51.4258**	1.211	2.4	ug/L	4	Standard
	Tl	205	438160.3	0.4	**54.6297**	1.400	2.6	ug/L	2	Standard
	Pb	206	306629.0	0.6	**51.1326**	1.396	2.7	ug/L	240	Standard
	Pb	207	272793.9	0.3	**51.2040**	1.495	2.9	ug/L	198	Standard
	Pb	208	1100211.2	0.6	**52.2716**	1.709	3.3	ug/L	883	Standard
	U	238	414469.5	0.5	**53.7134**	1.693	3.2	ug/L	3	Standard
>	Bi	209	587705.0	2.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 16:23:06
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**35.5299**	61.531	173.2	mg/L	0	Standard
Mg	24	4809.1	8.0	**5.4678**	0.353	6.5	mg/L	37	Standard
K	39	2090.1	1.6	**4.0082**	0.167	4.2	mg/L	38	Standard
Ca	43	111.7	10.3	**3.7837**	0.862	22.8	mg/L	10	Standard
Fe	54	7082.4	2.6	**5.2191**	0.217	4.1	mg/L	179	Standard
Fe	57	2040.1	4.3	**5.1108**	0.267	5.2	mg/L	335	Standard
Sc-1	45	68447.7	3.2				mg/L	61029	Standard
Cl	35	25830.7	1.4				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	2190.2	6.5				ug/L	1547	Standard
P	31	18406.2	0.9				ug/L	15286	Standard
S	34	1965.1	6.7				ug/L	1572	Standard
Sr	88	71.7	21.3				ug/L	73	Standard
C	12	103.3	29.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	152.5	102.6				mg/L	6	Standard
Ho-1	165	88.3	85.2				mg/L	10	Standard
Er	166	86.7	74.2				mg/L	13	Standard
I	127	2766.9	21.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	103.817		
Al	27	89.013		
Sc	45			
Ti	47	106.205		
V	51	110.094		
Cr	52	104.882		
Cr	53			
Mn	55	112.401		
Co	59	118.124		
Ni	60	104.355		
Cu	65	104.741		
Zn	66	105.936		
Ge	72		109.004	
As	75	104.876		
Se	82	109.798		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 16:23:06
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.479		
Ag	107	99.972		
Cd	111	98.481		
Cd	114			
In	115		116.255	
Sn	118	99.577		
Sb	123	97.186		
Ba	135	93.802		
Ce	140			
Tb	159			
Ho	165			
Tl	203	102.852		
Tl	205			
Pb	206			
Pb	207			
Pb	208	104.543		
U	238	107.427		
Bi	209		112.161	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	V	51	
QC Std 6	Mn	55	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 16:23:06
Page 3

Approved: July 04, 2016

QC Std 6	Co	59

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 16:24:01
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100067.7	1.3				ug/L	74699	Standard
	Be	9	36.7	7.9	**0.0387**	0.002	5.2	ug/L	8	Standard
	Al	27	10153.6	30.4	**0.0381**	0.015	39.5	ug/L	1493	Standard
	Sc	45	67434.8	2.7				ug/L	61029	Standard
	Ti	47	38.3	21.7	**-0.0212**	0.020	92.7	ug/L	35	Standard
	V	51	1502.7	1.2	**-0.0231**	0.006	25.9	ug/L	1435	Standard
	Cr	52	7451.2	2.4	**-0.1523**	0.012	7.6	ug/L	7635	Standard
	Cr	53	503.3	11.6	**0.1429**	0.063	44.4	ug/L	363	Standard
	Mn	55	3693.5	26.8	**0.3137**	0.134	42.8	ug/L	747	Standard
	Co	59	358.3	10.7	**0.0099**	0.006	62.4	ug/L	234	Standard
	Ni	60	209.7	4.9	**-0.0268**	0.006	22.7	ug/L	215	Standard
	Cu	65	198.7	7.2	**-0.0110**	0.009	80.7	ug/L	104	Standard
	Zn	66	532.3	5.7	**-0.0316**	0.034	106.3	ug/L	363	Standard
>	**Ge**	72	571298.3	2.3				ug/L	495142	Standard
	As	75	-12.4	42.9	**0.0499**	0.004	8.8	ug/L	-50	Standard
	Se	82	23.5	9.9	**0.0476**	0.021	43.8	ug/L	14	Standard
	Se-1	77	58.7	7.7	**0.0749**	0.069	91.8	ug/L	42	Standard
>	**Ga**	71	75.0	26.7				mg/L	42	Standard
	Rb	85	36.7	55.1				ug/L	17	Standard
	Y	89	500557.3	2.5				ug/L	417098	Standard
>	**Rh**	103	30.0	44.1				ug/L	13	Standard
	Mo	98	301.0	26.9	**0.0924**	0.026	27.9	ug/L	11	Standard
	Ag	107	165.0	21.0	**0.0091**	0.004	49.2	ug/L	68	Standard
	Cd	111	19.8	28.3	**0.0003**	0.002	699.6	mg/L	2	Standard
	Cd	114	46.5	85.8	**0.0058**	0.007	116.4	ug/L	12	Standard
>	**In**	115	675355.4	0.8				ug/L	587441	Standard
	Sn	118	876.7	18.6	**0.0359**	0.025	69.7	ug/L	497	Standard
	Sb	123	599.5	22.8	**0.1076**	0.026	24.3	ug/L	78	Standard
	Ba	135	42.7	23.9	**0.0021**	0.004	201.6	ug/L	24	Standard
	Ce	140	140.0	34.4				ug/L	25	Standard
>	**Tb**	159	1044716.9	1.2				ug/L	939701	Standard
	Ho	165	20.0	50.0				ug/L	10	Standard
	Tl	203	103.0	21.9	**0.0094**	0.003	27.0	ug/L	4	Standard
	Tl	205	70.0	58.5	**0.0089**	0.005	57.1	ug/L	2	Standard
	Pb	206	338.7	4.4	**-0.0148**	0.002	16.0	ug/L	240	Standard
	Pb	207	277.7	9.1	**-0.0179**	0.005	28.1	ug/L	198	Standard
	Pb	208	1213.0	9.6	**-0.0165**	0.006	36.1	ug/L	883	Standard
	U	238	62.0	33.9	**0.0036**	0.003	76.1	ug/L	3	Standard
>	**Bi**	209	588682.5	1.2				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 16:26:17
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.0278**	29.484	173.2	mg/L	0	Standard
Mg	24	56.7	20.4	**0.0014**	0.015	1120.4	mg/L	37	Standard
K	39	36.7	28.4	**-0.0407**	0.021	51.2	mg/L	38	Standard
Ca	43	26.7	28.6	**-1.0029**	0.415	41.4	mg/L	10	Standard
Fe	54	187.4	12.0	**0.0055**	0.017	304.1	mg/L	179	Standard
Fe	57	261.7	6.1	**-0.2681**	0.071	26.5	mg/L	335	Standard
Sc-1	45	67434.8	2.7				mg/L	61029	Standard
Cl	35	26219.3	0.7				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	2243.5	3.0				ug/L	1547	Standard
P	31	19277.3	3.9				ug/L	15286	Standard
S	34	2016.8	11.3				ug/L	1572	Standard
Sr	88	70.0	35.7				ug/L	73	Standard
C	12	120.0	25.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	49.4	87.7				mg/L	6	Standard
Ho-1	165	20.0	50.0				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	2125.2	22.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.381	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 16:26:17
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.966
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		112.347
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 16:26:17
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701SDL WG574688-02
Sample Date/Time: Wednesday, June 29, 2016 16:27:13
Number of Replicates: 3
Autosampler Position: 215
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	105717.5	0.5				ug/L	74699	Standard
	Be	9	553.3	1.4	**0.4517**	0.004	0.8	ug/L	8	Standard
	Al	27	5662006.5	2.5	**27.2120**	0.787	2.9	ug/L	1493	Standard
	Sc	45	72131.5	3.8				ug/L	61029	Standard
	Ti	47	604.0	7.1	**1.2783**	0.109	8.5	ug/L	35	Standard
	V	51	3941.0	1.4	**0.2585**	0.002	0.9	ug/L	1435	Standard
	Cr	52	12715.3	2.3	**0.4691**	0.023	4.9	ug/L	7635	Standard
	Cr	53	1340.1	9.1	**0.9313**	0.103	11.1	ug/L	363	Standard
	Mn	55	2452895.1	0.8	**306.4368**	3.099	1.0	ug/L	747	Standard
	Co	59	67057.1	0.2	**8.7998**	0.112	1.3	ug/L	234	Standard
	Ni	60	46732.3	0.4	**21.5970**	0.233	1.1	ug/L	215	Standard
	Cu	65	904.0	6.3	**0.3188**	0.021	6.7	ug/L	104	Standard
	Zn	66	62660.6	1.1	**49.4618**	0.122	0.2	ug/L	363	Standard
>	**Ge**	72	569030.6	1.4				ug/L	495142	Standard
	As	75	59.1	19.2	**0.1055**	0.009	8.4	ug/L	-50	Standard
	Se	82	34.4	3.8	**0.1350**	0.014	10.1	ug/L	14	Standard
	Se-1	77	82.7	11.5	**0.3621**	0.106	29.4	ug/L	42	Standard
>	**Ga**	71	288.3	19.2				mg/L	42	Standard
	Rb	85	7325.1	4.9				ug/L	17	Standard
	Y	89	616528.9	0.8				ug/L	417098	Standard
>	**Rh**	103	13.3	78.1				ug/L	13	Standard
	Mo	98	112.4	8.7	**0.0298**	0.003	9.5	ug/L	11	Standard
	Ag	107	122.3	20.4	**0.0029**	0.003	108.6	ug/L	68	Standard
	Cd	111	198.8	14.6	**0.0723**	0.012	16.2	mg/L	2	Standard
	Cd	114	516.3	5.9	**0.0829**	0.006	6.9	ug/L	12	Standard
>	**In**	115	696026.2	0.9				ug/L	587441	Standard
	Sn	118	760.0	4.7	**0.0151**	0.006	37.2	ug/L	497	Standard
	Sb	123	350.2	21.1	**0.0592**	0.014	23.4	ug/L	78	Standard
	Ba	135	775.0	6.3	**0.3035**	0.019	6.3	ug/L	24	Standard
	Ce	140	119180.2	0.3				ug/L	25	Standard
>	**Tb**	159	1103880.2	1.1				ug/L	939701	Standard
	Ho	165	9261.2	2.5				ug/L	10	Standard
	Tl	203	155.7	9.5	**0.0145**	0.001	9.9	ug/L	4	Standard
	Tl	205	156.7	24.4	**0.0191**	0.005	23.7	ug/L	2	Standard
	Pb	206	668.3	5.9	**0.0368**	0.005	14.8	ug/L	240	Standard
	Pb	207	530.0	3.8	**0.0264**	0.004	13.6	ug/L	198	Standard
	Pb	208	2281.7	2.6	**0.0309**	0.002	6.9	ug/L	883	Standard
	U	238	1160.7	2.8	**0.1411**	0.004	2.6	ug/L	3	Standard
>	**Bi**	209	607087.1	0.9				ug/L	523984	Standard

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:29:30
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.5008**	28.572	173.2	mg/L	0	Standard
Mg	24	3728.8	3.6	**4.0090**	0.135	3.4	mg/L	37	Standard
K	39	85.0	42.4	**0.0464**	0.073	156.7	mg/L	38	Standard
Ca	43	183.3	11.4	**7.2814**	1.162	16.0	mg/L	10	Standard
Fe	54	8713.3	2.8	**6.1146**	0.148	2.4	mg/L	179	Standard
Fe	57	2506.9	1.7	**6.1424**	0.321	5.2	mg/L	335	Standard
> **Sc-1**	45	72131.5	3.8				mg/L	61029	Standard
Cl	35	26208.6	1.1				ug/L	20195	Standard
Kr	83	2.7	21.7				ug/L	4	Standard
Br	81	3227.0	4.8				ug/L	1547	Standard
P	31	20363.7	0.5				ug/L	15286	Standard
S	34	1863.4	4.5				ug/L	1572	Standard
Sr	88	105.0	21.8				ug/L	73	Standard
C	12	56.7	36.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	14855.4	4.1				mg/L	6	Standard
Ho-1	165	9261.2	2.5				mg/L	10	Standard
Er	166	8048.8	4.7				mg/L	13	Standard
I	127	5796.1	1.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		141.525	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.923	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:29:30
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	118.484
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	115.860
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:29:30
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126504
Sample Date/Time: Wednesday, June 29, 2016 16:30:25
Number of Replicates: 3
Autosampler Position: 216
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101552.6	1.9				ug/L	74699	Standard
	Be	9	25.0	72.1	**0.0287**	0.016	54.3	ug/L	8	Standard
	Al	27	16250197.0	1.8	**81.3578**	2.887	3.5	ug/L	1493	Standard
	Sc	45	67044.7	1.9				ug/L	61029	Standard
	Ti	47	394.3	10.9	**0.8437**	0.094	11.2	ug/L	35	Standard
	V	51	39093.2	1.3	**4.5488**	0.169	3.7	ug/L	1435	Standard
	Cr	52	19068.3	2.3	**1.3348**	0.101	7.6	ug/L	7635	Standard
	Cr	53	5326.0	20.8	**4.9620**	1.165	23.5	ug/L	363	Standard
	Mn	55	6239.0	2.5	**0.6731**	0.027	4.0	ug/L	747	Standard
	Co	59	851.0	2.5	**0.0808**	0.004	4.5	ug/L	234	Standard
	Ni	60	1951.8	2.1	**0.8314**	0.029	3.5	ug/L	215	Standard
	Cu	65	1321.1	4.3	**0.5464**	0.014	2.5	ug/L	104	Standard
	Zn	66	5985.8	0.6	**4.5673**	0.118	2.6	ug/L	363	Standard
>	**Ge**	72	540512.2	2.5				ug/L	495142	Standard
	As	75	1446.2	5.5	**1.2419**	0.041	3.3	ug/L	-50	Standard
	Se	82	841.6	2.6	**6.8893**	0.246	3.6	ug/L	14	Standard
	Se-1	77	609.0	15.5	**7.0254**	1.243	17.7	ug/L	42	Standard
>	**Ga**	71	105.0	12.6				mg/L	42	Standard
	Rb	85	1541.7	7.2				ug/L	17	Standard
	Y	89	489017.5	2.4				ug/L	417098	Standard
>	**Rh**	103	70.0	18.9				ug/L	13	Standard
	Mo	98	501.8	4.3	**0.1636**	0.005	2.9	ug/L	11	Standard
	Ag	107	97.0	7.4	**0.0005**	0.001	200.8	ug/L	68	Standard
	Cd	111	26.4	23.0	**0.0035**	0.003	79.3	mg/L	2	Standard
	Cd	114	74.1	18.4	**0.0109**	0.002	21.9	ug/L	12	Standard
>	**In**	115	652860.7	1.6				ug/L	587441	Standard
	Sn	118	840.0	9.7	**0.0347**	0.012	35.4	ug/L	497	Standard
	Sb	123	289.8	10.9	**0.0517**	0.007	12.7	ug/L	78	Standard
	Ba	135	17710.3	0.6	**7.7715**	0.161	2.1	ug/L	24	Standard
	Ce	140	1525.1	0.6				ug/L	25	Standard
>	**Tb**	159	1031518.0	1.9				ug/L	939701	Standard
	Ho	165	43.3	24.0				ug/L	10	Standard
	Tl	203	163.3	10.6	**0.0174**	0.002	9.6	ug/L	4	Standard
	Tl	205	133.3	16.9	**0.0182**	0.003	18.7	ug/L	2	Standard
	Pb	206	441.3	2.3	**0.0087**	0.001	8.9	ug/L	240	Standard
	Pb	207	362.0	4.8	**0.0037**	0.003	70.2	ug/L	198	Standard
	Pb	208	1549.4	2.6	**0.0058**	0.001	11.8	ug/L	883	Standard
	U	238	23538.3	0.9	**3.3015**	0.086	2.6	ug/L	3	Standard
>	**Bi**	209	542188.4	2.1				ug/L	523984	Standard

Sample ID: L1606126504
Report Date/Time: Wednesday, June 29, 2016 16:32:42
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**35.7447**	61.903	173.2	mg/L	0	Standard
Mg	24	119069.4	0.9	**139.8814**	3.117	2.2	mg/L	37	Standard
K	39	143.3	11.2	**0.1739**	0.028	15.9	mg/L	38	Standard
Ca	43	463.3	1.6	**24.1242**	0.657	2.7	mg/L	10	Standard
Fe	54	197.6	4.4	**0.0142**	0.006	45.1	mg/L	179	Standard
Fe	57	435.0	13.1	**0.2738**	0.196	71.7	mg/L	335	Standard
> **Sc-1**	45	67044.7	1.9				mg/L	61029	Standard
Cl	35	31470.7	2.7				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	73323.9	4.5				ug/L	1547	Standard
P	31	23608.4	2.5				ug/L	15286	Standard
S	34	1800.1	6.6				ug/L	1572	Standard
Sr	88	293.3	6.9				ug/L	73	Standard
C	12	193.3	41.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	86.5	19.9				mg/L	6	Standard
Ho-1	165	43.3	24.0				mg/L	10	Standard
Er	166	73.3	34.3				mg/L	13	Standard
I	127	54762.3	5.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		135.949	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.163	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126504
Report Date/Time: Wednesday, June 29, 2016 16:32:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.136
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.474
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Sample ID: L1606126504
Report Date/Time: Wednesday, June 29, 2016 16:32:42
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126505
Sample Date/Time: Wednesday, June 29, 2016 16:33:36
Number of Replicates: 3
Autosampler Position: 217
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	99450.4	1.4				ug/L	74699	Standard
	Be	9	58.3	30.1	**0.0575**	0.016	27.1	ug/L	8	Standard
	Al	27	32118811.5	1.4	**164.1461**	0.108	0.1	ug/L	1493	Standard
	Sc	45	70749.8	1.3				ug/L	61029	Standard
	Ti	47	642.0	2.2	**1.4252**	0.030	2.1	ug/L	35	Standard
	V	51	8228.7	11.8	**0.7915**	0.119	15.1	ug/L	1435	Standard
	Cr	52	11308.2	2.8	**0.3584**	0.044	12.4	ug/L	7635	Standard
	Cr	53	3712.1	10.4	**3.3091**	0.391	11.8	ug/L	363	Standard
	Mn	55	3368829.0	1.1	**438.1329**	6.772	1.5	ug/L	747	Standard
	Co	59	3400.7	28.1	**0.4294**	0.132	30.8	ug/L	234	Standard
	Ni	60	4538.7	3.1	**2.0720**	0.077	3.7	ug/L	215	Standard
	Cu	65	1589.1	4.3	**0.6697**	0.036	5.3	ug/L	104	Standard
	Zn	66	6182.6	5.7	**4.6729**	0.311	6.7	ug/L	363	Standard
>	**Ge**	72	546665.5	0.5				ug/L	495142	Standard
	As	75	6156.5	0.5	**5.0379**	0.006	0.1	ug/L	-50	Standard
	Se	82	164.0	1.8	**1.2151**	0.026	2.2	ug/L	14	Standard
	Se-1	77	258.7	9.7	**2.5862**	0.322	12.5	ug/L	42	Standard
>	**Ga**	71	188.3	13.1				mg/L	42	Standard
	Rb	85	1378.4	10.6				ug/L	17	Standard
	Y	89	484009.6	2.1				ug/L	417098	Standard
>	**Rh**	103	118.3	6.5				ug/L	13	Standard
	Mo	98	286.3	31.1	**0.0925**	0.030	32.6	ug/L	11	Standard
	Ag	107	183.3	43.5	**0.0128**	0.011	86.7	ug/L	68	Standard
	Cd	111	248.2	17.8	**0.1006**	0.018	18.1	mg/L	2	Standard
	Cd	114	590.9	17.2	**0.1032**	0.017	16.3	ug/L	12	Standard
>	**In**	115	642069.8	1.9				ug/L	587441	Standard
	Sn	118	836.7	5.6	**0.0365**	0.010	26.8	ug/L	497	Standard
	Sb	123	288.3	29.8	**0.0524**	0.017	32.5	ug/L	78	Standard
	Ba	135	131296.8	1.0	**58.6819**	0.595	1.0	ug/L	24	Standard
	Ce	140	6513.1	10.5				ug/L	25	Standard
>	**Tb**	159	1012113.9	1.4				ug/L	939701	Standard
	Ho	165	216.7	7.1				ug/L	10	Standard
	Tl	203	185.3	61.4	**0.0202**	0.013	66.8	ug/L	4	Standard
	Tl	205	276.7	88.2	**0.0379**	0.033	87.8	ug/L	2	Standard
	Pb	206	891.4	31.8	**0.0916**	0.052	56.6	ug/L	240	Standard
	Pb	207	748.7	30.8	**0.0839**	0.048	56.6	ug/L	198	Standard
	Pb	208	3050.1	30.7	**0.0846**	0.049	57.6	ug/L	883	Standard
	U	238	21997.6	2.6	**3.1135**	0.091	2.9	ug/L	3	Standard
>	**Bi**	209	537104.2	0.4				ug/L	523984	Standard

Sample ID: L1606126505
Report Date/Time: Wednesday, June 29, 2016 16:35:53
Page 1

Approved: July 04, 2016

Na	23	13.3	57.3	**132.8617**	77.017	58.0	mg/L	0	Standard
Mg	24	104917.1	1.8	**116.7540**	0.597	0.5	mg/L	37	Standard
K	39	116.7	8.9	**0.1082**	0.018	16.4	mg/L	38	Standard
Ca	43	916.7	6.0	**47.4532**	3.610	7.6	mg/L	10	Standard
Fe	54	3013.9	2.4	**2.0666**	0.083	4.0	mg/L	179	Standard
Fe	57	1381.7	4.1	**2.9770**	0.162	5.4	mg/L	335	Standard
> **Sc-1**	45	70749.8	1.3				mg/L	61029	Standard
Cl	35	30215.4	1.0				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	35768.9	2.6				ug/L	1547	Standard
P	31	25429.7	2.5				ug/L	15286	Standard
S	34	2313.5	0.3				ug/L	1572	Standard
Sr	88	551.7	8.2				ug/L	73	Standard
C	12	433.3	25.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	3	Standard
Dy	164	279.9	12.3				mg/L	6	Standard
Ho-1	165	216.7	7.1				mg/L	10	Standard
Er	166	280.0	18.9				mg/L	13	Standard
I	127	794211.5	6.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		133.135	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.406	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126505
Report Date/Time: Wednesday, June 29, 2016 16:35:53
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.299
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.504
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606126505
Report Date/Time: Wednesday, June 29, 2016 16:35:53
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126506
Sample Date/Time: Wednesday, June 29, 2016 16:36:48
Number of Replicates: 3
Autosampler Position: 218
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100252.3	1.1				ug/L	74699	Standard
	Be	9	26.7	39.0	**0.0303**	0.009	29.8	ug/L	8	Standard
	Al	27	69444.0	4.9	**0.3387**	0.017	5.1	ug/L	1493	Standard
	Sc	45	71205.3	1.8				ug/L	61029	Standard
	Ti	47	113.0	20.9	**0.1414**	0.055	38.5	ug/L	35	Standard
	V	51	2901.2	2.9	**0.1267**	0.013	10.0	ug/L	1435	Standard
	Cr	52	23756.9	1.1	**1.6651**	0.003	0.2	ug/L	7635	Standard
	Cr	53	2710.2	4.1	**2.1308**	0.127	5.9	ug/L	363	Standard
	Mn	55	7460.8	10.3	**0.7511**	0.098	13.0	ug/L	747	Standard
	Co	59	456.0	8.6	**0.0207**	0.006	26.8	ug/L	234	Standard
	Ni	60	2497.2	2.2	**0.9949**	0.023	2.3	ug/L	215	Standard
	Cu	65	984.0	1.2	**0.3398**	0.007	2.0	ug/L	104	Standard
	Zn	66	8303.3	0.9	**5.9218**	0.058	1.0	ug/L	363	Standard
>	**Ge**	72	590298.4	1.1				ug/L	495142	Standard
	As	75	7.9	237.8	**0.0655**	0.014	21.6	ug/L	-50	Standard
	Se	82	27.7	14.2	**0.0742**	0.032	43.2	ug/L	14	Standard
	Se-1	77	72.3	15.2	**0.2091**	0.134	63.9	ug/L	42	Standard
>	**Ga**	71	91.7	20.7				mg/L	42	Standard
	Rb	85	308.3	4.1				ug/L	17	Standard
	Y	89	498995.2	1.7				ug/L	417098	Standard
>	**Rh**	103	21.7	48.0				ug/L	13	Standard
	Mo	98	716.7	9.5	**0.2259**	0.023	10.1	ug/L	11	Standard
	Ag	107	155.3	16.4	**0.0076**	0.003	44.6	ug/L	68	Standard
	Cd	111	44.4	12.9	**0.0104**	0.002	21.8	mg/L	2	Standard
	Cd	114	117.5	18.7	**0.0177**	0.004	21.5	ug/L	12	Standard
>	**In**	115	682064.0	0.6				ug/L	587441	Standard
	Sn	118	4689.1	2.4	**0.5966**	0.013	2.3	ug/L	497	Standard
	Sb	123	228.7	11.3	**0.0380**	0.005	12.5	ug/L	78	Standard
	Ba	135	1034.0	4.7	**0.4191**	0.020	4.8	ug/L	24	Standard
	Ce	140	486.7	14.8				ug/L	25	Standard
>	**Tb**	159	1056147.0	0.9				ug/L	939701	Standard
	Ho	165	8.3	69.3				ug/L	10	Standard
	Tl	203	43.7	26.0	**0.0029**	0.001	43.8	ug/L	4	Standard
	Tl	205	50.0	26.5	**0.0063**	0.002	26.5	ug/L	2	Standard
	Pb	206	665.7	3.1	**0.0376**	0.003	7.4	ug/L	240	Standard
	Pb	207	542.7	5.2	**0.0299**	0.006	20.6	ug/L	198	Standard
	Pb	208	2227.4	4.5	**0.0297**	0.006	20.4	ug/L	883	Standard
	U	238	38.7	20.9	**0.0005**	0.001	221.1	ug/L	3	Standard
>	**Bi**	209	599996.8	1.5				ug/L	523984	Standard

Sample ID: L1606126506
Report Date/Time: Wednesday, June 29, 2016 16:39:04
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	216.7	5.8	**0.1745**	0.010	5.5	mg/L	37	Standard
K	39	46.7	49.5	**-0.0257**	0.044	170.8	mg/L	38	Standard
Ca	43	30.0	16.7	**-0.8971**	0.290	32.3	mg/L	10	Standard
Fe	54	200.7	16.8	**0.0076**	0.025	326.7	mg/L	179	Standard
Fe	57	248.3	11.1	**-0.3516**	0.067	19.2	mg/L	335	Standard
> **Sc-1**	45	71205.3	1.8				mg/L	61029	Standard
Cl	35	31458.6	1.1				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	3110.3	10.8				ug/L	1547	Standard
P	31	21037.9	1.6				ug/L	15286	Standard
S	34	2263.5	3.0				ug/L	1572	Standard
Sr	88	58.3	27.6				ug/L	73	Standard
C	12	113.3	13.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	38.9	27.0				mg/L	6	Standard
Ho-1	165	8.3	69.3				mg/L	10	Standard
Er	166	23.3	65.5				mg/L	13	Standard
I	127	20210.2	61.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		134.209	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		119.218	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126506
Report Date/Time: Wednesday, June 29, 2016 16:39:04
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.108
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		114.507
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126601
Sample Date/Time: Wednesday, June 29, 2016 16:39:59
Number of Replicates: 3
Autosampler Position: 219
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101321.2	3.1				ug/L	74699	Standard
	Be	9	18.3	56.8	**0.0228**	0.008	36.0	ug/L	8	Standard
	Al	27	44431569.1	0.7	**222.9926**	5.415	2.4	ug/L	1493	Standard
	Sc	45	71310.8	2.2				ug/L	61029	Standard
	Ti	47	275.3	6.3	**0.5272**	0.047	8.9	ug/L	35	Standard
	V	51	27907.2	1.9	**3.0387**	0.030	1.0	ug/L	1435	Standard
	Cr	52	13580.1	0.8	**0.5803**	0.015	2.6	ug/L	7635	Standard
	Cr	53	2821.9	2.8	**2.3430**	0.081	3.4	ug/L	363	Standard
	Mn	55	153037.1	1.2	**19.0926**	0.078	0.4	ug/L	747	Standard
	Co	59	1012.0	1.5	**0.0969**	0.001	1.5	ug/L	234	Standard
	Ni	60	4370.3	2.4	**1.9194**	0.048	2.5	ug/L	215	Standard
	Cu	65	751.7	5.3	**0.2499**	0.020	8.0	ug/L	104	Standard
	Zn	66	3518.1	1.4	**2.3650**	0.050	2.1	ug/L	363	Standard
>	**Ge**	72	565652.8	1.1				ug/L	495142	Standard
	As	75	1269.0	4.3	**1.0512**	0.035	3.4	ug/L	-50	Standard
	Se	82	382.5	3.3	**2.9130**	0.125	4.3	ug/L	14	Standard
	Se-1	77	193.3	14.8	**1.6975**	0.369	21.7	ug/L	42	Standard
>	**Ga**	71	66.7	18.9				mg/L	42	Standard
	Rb	85	2323.5	7.7				ug/L	17	Standard
	Y	89	480038.2	1.2				ug/L	417098	Standard
>	**Rh**	103	150.0	3.3				ug/L	13	Standard
	Mo	98	638.7	4.4	**0.2084**	0.010	5.0	ug/L	11	Standard
	Ag	107	92.0	14.1	**-0.0003**	0.002	569.4	ug/L	68	Standard
	Cd	111	195.2	3.6	**0.0755**	0.003	3.5	mg/L	2	Standard
	Cd	114	469.8	19.5	**0.0797**	0.016	19.7	ug/L	12	Standard
>	**In**	115	657497.5	0.6				ug/L	587441	Standard
	Sn	118	753.4	8.5	**0.0205**	0.010	48.9	ug/L	497	Standard
	Sb	123	222.8	13.6	**0.0384**	0.006	15.3	ug/L	78	Standard
	Ba	135	100054.6	0.7	**43.6617**	0.503	1.2	ug/L	24	Standard
	Ce	140	173.3	6.0				ug/L	25	Standard
>	**Tb**	159	1045560.4	1.5				ug/L	939701	Standard
	Ho	165	40.0	25.0				ug/L	10	Standard
	Tl	203	129.3	1.6	**0.0127**	0.000	3.7	ug/L	4	Standard
	Tl	205	133.3	20.7	**0.0174**	0.003	20.0	ug/L	2	Standard
	Pb	206	412.3	3.7	**0.0001**	0.002	2456.7	ug/L	240	Standard
	Pb	207	343.7	8.9	**-0.0032**	0.005	165.7	ug/L	198	Standard
	Pb	208	1432.7	1.3	**-0.0035**	0.002	56.3	ug/L	883	Standard
	U	238	43636.7	0.7	**5.8505**	0.120	2.1	ug/L	3	Standard
>	**Bi**	209	567479.2	1.8				ug/L	523984	Standard

Sample ID: L1606126601
Report Date/Time: Wednesday, June 29, 2016 16:42:16
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**48.6438**	48.507	99.7	mg/L	0	Standard
Mg	24	35550.0	1.9	**39.2140**	0.699	1.8	mg/L	37	Standard
K	39	165.0	24.1	**0.1975**	0.071	36.0	mg/L	38	Standard
Ca	43	596.7	5.4	**29.7553**	2.186	7.3	mg/L	10	Standard
Fe	54	190.1	10.6	**-0.0004**	0.014	3403.7	mg/L	179	Standard
Fe	57	555.0	7.7	**0.5391**	0.093	17.2	mg/L	335	Standard
> **Sc-1**	45	71310.8	2.2				mg/L	61029	Standard
Cl	35	33374.1	0.9				ug/L	20195	Standard
Kr	83	2.0	132.3				ug/L	4	Standard
Br	81	77209.6	1.2				ug/L	1547	Standard
P	31	24828.7	1.5				ug/L	15286	Standard
S	34	2263.5	4.7				ug/L	1572	Standard
Sr	88	591.7	14.4				ug/L	73	Standard
C	12	253.3	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	40.0	25.0				mg/L	3	Standard
Dy	164	40.1	59.8				mg/L	6	Standard
Ho-1	165	40.0	25.0				mg/L	10	Standard
Er	166	66.7	37.7				mg/L	13	Standard
I	127	2058609.0	7.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		135.640	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.240	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126601
Report Date/Time: Wednesday, June 29, 2016 16:42:16
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.926
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.301
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126901
Sample Date/Time: Wednesday, June 29, 2016 16:43:10
Number of Replicates: 3
Autosampler Position: 220
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	97179.0	1.2				ug/L	74699	Standard
	Be	9	23.3	44.6	**0.0281**	0.009	33.4	ug/L	8	Standard
	Al	27	25054309.4	1.7	**131.0332**	1.604	1.2	ug/L	1493	Standard
	Sc	45	69363.5	2.7				ug/L	61029	Standard
	Ti	47	323.3	2.8	**0.6992**	0.021	2.9	ug/L	35	Standard
	V	51	12024.0	2.6	**1.3128**	0.037	2.9	ug/L	1435	Standard
	Cr	52	11840.6	1.6	**0.4900**	0.025	5.0	ug/L	7635	Standard
	Cr	53	2901.9	3.9	**2.6451**	0.119	4.5	ug/L	363	Standard
	Mn	55	490032.6	0.8	**66.5395**	0.816	1.2	ug/L	747	Standard
	Co	59	931.7	2.1	**0.0964**	0.003	3.3	ug/L	234	Standard
	Ni	60	2217.5	3.2	**0.9981**	0.030	3.0	ug/L	215	Standard
	Cu	65	599.0	3.0	**0.2013**	0.011	5.4	ug/L	104	Standard
	Zn	66	4806.1	1.2	**3.7141**	0.039	1.0	ug/L	363	Standard
>	**Ge**	72	522601.4	0.6				ug/L	495142	Standard
	As	75	2361.3	3.3	**2.0571**	0.073	3.5	ug/L	-50	Standard
	Se	82	593.9	4.2	**4.9896**	0.230	4.6	ug/L	14	Standard
	Se-1	77	179.7	2.8	**1.7086**	0.076	4.5	ug/L	42	Standard
>	**Ga**	71	105.0	9.5				mg/L	42	Standard
	Rb	85	2805.3	2.5				ug/L	17	Standard
	Y	89	474068.5	3.1				ug/L	417098	Standard
>	**Rh**	103	81.7	12.7				ug/L	13	Standard
	Mo	98	305.2	1.9	**0.0985**	0.004	4.4	ug/L	11	Standard
	Ag	107	99.7	12.7	**0.0010**	0.002	153.4	ug/L	68	Standard
	Cd	111	19.2	16.0	**0.0004**	0.001	278.0	mg/L	2	Standard
	Cd	114	49.5	42.7	**0.0067**	0.004	59.0	ug/L	12	Standard
>	**In**	115	645532.6	2.3				ug/L	587441	Standard
	Sn	118	888.4	9.9	**0.0435**	0.010	24.0	ug/L	497	Standard
	Sb	123	240.8	4.4	**0.0427**	0.002	4.3	ug/L	78	Standard
	Ba	135	155622.1	0.5	**69.2020**	1.756	2.5	ug/L	24	Standard
	Ce	140	928.4	7.9				ug/L	25	Standard
>	**Tb**	159	1004174.2	1.0				ug/L	939701	Standard
	Ho	165	23.3	32.7				ug/L	10	Standard
	Tl	203	115.7	10.5	**0.0121**	0.001	11.5	ug/L	4	Standard
	Tl	205	128.3	22.2	**0.0179**	0.004	22.0	ug/L	2	Standard
	Pb	206	395.0	0.9	**0.0016**	0.001	57.1	ug/L	240	Standard
	Pb	207	323.3	6.0	**-0.0029**	0.004	138.0	ug/L	198	Standard
	Pb	208	1415.7	2.0	**0.0002**	0.001	432.7	ug/L	883	Standard
	U	238	7265.7	2.0	**1.0348**	0.023	2.2	ug/L	3	Standard
>	**Bi**	209	532226.3	0.6				ug/L	523984	Standard

Sample ID: L1606126901
Report Date/Time: Wednesday, June 29, 2016 16:45:27
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.4177**	30.160	173.2	mg/L	0	Standard
Mg	24	119202.1	1.4	**135.3768**	3.695	2.7	mg/L	37	Standard
K	39	220.0	8.2	**0.3143**	0.042	13.4	mg/L	38	Standard
Ca	43	398.3	4.4	**19.6130**	0.698	3.6	mg/L	10	Standard
Fe	54	252.1	18.1	**0.0493**	0.029	59.6	mg/L	179	Standard
Fe	57	418.3	9.7	**0.1785**	0.138	77.4	mg/L	335	Standard
Sc-1	45	69363.5	2.7				mg/L	61029	Standard
Cl	35	32454.7	1.8				ug/L	20195	Standard
Kr	83	1.0	173.2				ug/L	4	Standard
Br	81	137582.4	3.7				ug/L	1547	Standard
P	31	24828.7	1.5				ug/L	15286	Standard
S	34	2306.8	3.7				ug/L	1572	Standard
Sr	88	291.7	6.0				ug/L	73	Standard
C	12	323.3	19.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	15.2	73.1				mg/L	6	Standard
Ho-1	165	23.3	32.7				mg/L	10	Standard
Er	166	30.0	33.3				mg/L	13	Standard
I	127	310633.7	2.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		130.094	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		105.546	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606126901
Report Date/Time: Wednesday, June 29, 2016 16:45:27
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.889
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.573
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127201
Sample Date/Time: Wednesday, June 29, 2016 16:46:22
Number of Replicates: 3
Autosampler Position: 221
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94337.9	2.3				ug/L	74699	Standard
	Be	9	23.3	53.9	**0.0288**	0.012	40.8	ug/L	8	Standard
	Al	27	2978340.1	2.2	**16.0395**	0.522	3.3	ug/L	1493	Standard
	Sc	45	67137.0	4.7				ug/L	61029	Standard
	Ti	47	236.7	18.3	**0.4520**	0.099	21.9	ug/L	35	Standard
	V	51	33734.3	0.8	**3.8201**	0.030	0.8	ug/L	1435	Standard
	Cr	52	13200.4	3.0	**0.5779**	0.062	10.7	ug/L	7635	Standard
	Cr	53	1396.7	5.5	**1.0289**	0.077	7.5	ug/L	363	Standard
	Mn	55	5810.0	55.0	**0.6034**	0.416	68.9	ug/L	747	Standard
	Co	59	597.0	23.2	**0.0441**	0.019	43.4	ug/L	234	Standard
	Ni	60	839.7	12.2	**0.2798**	0.055	19.7	ug/L	215	Standard
	Cu	65	499.7	4.1	**0.1378**	0.013	9.3	ug/L	104	Standard
	Zn	66	1881.8	10.3	**1.0958**	0.181	16.5	ug/L	363	Standard
>	**Ge**	72	550779.1	1.5				ug/L	495142	Standard
	As	75	823.5	3.7	**0.7205**	0.021	2.9	ug/L	-50	Standard
	Se	82	534.3	3.8	**4.2403**	0.211	5.0	ug/L	14	Standard
	Se-1	77	394.7	5.2	**4.2338**	0.183	4.3	ug/L	42	Standard
>	**Ga**	71	76.7	13.6				mg/L	42	Standard
	Rb	85	1120.0	6.4				ug/L	17	Standard
	Y	89	478775.5	2.0				ug/L	417098	Standard
>	**Rh**	103	33.3	52.7				ug/L	13	Standard
	Mo	98	1510.8	4.0	**0.5152**	0.028	5.4	ug/L	11	Standard
	Ag	107	155.0	42.6	**0.0091**	0.010	106.9	ug/L	68	Standard
	Cd	111	66.6	64.6	**0.0216**	0.019	89.7	mg/L	2	Standard
	Cd	114	188.5	50.4	**0.0319**	0.018	55.0	ug/L	12	Standard
>	**In**	115	639482.4	2.0				ug/L	587441	Standard
	Sn	118	738.4	13.9	**0.0215**	0.017	79.5	ug/L	497	Standard
	Sb	123	301.8	22.4	**0.0554**	0.014	26.0	ug/L	78	Standard
	Ba	135	10964.6	1.7	**4.9054**	0.035	0.7	ug/L	24	Standard
	Ce	140	728.4	65.5				ug/L	25	Standard
>	**Tb**	159	995838.4	2.1				ug/L	939701	Standard
	Ho	165	55.0	78.7				ug/L	10	Standard
	Tl	203	165.3	64.1	**0.0175**	0.012	69.7	ug/L	4	Standard
	Tl	205	88.3	34.1	**0.0121**	0.004	34.3	ug/L	2	Standard
	Pb	206	371.3	17.5	**-0.0043**	0.012	275.3	ug/L	240	Standard
	Pb	207	297.3	13.9	**-0.0098**	0.008	82.9	ug/L	198	Standard
	Pb	208	1253.7	10.6	**-0.0098**	0.007	68.1	ug/L	883	Standard
	U	238	7715.3	1.4	**1.0741**	0.027	2.5	ug/L	3	Standard
>	**Bi**	209	544657.1	1.1				ug/L	523984	Standard

Sample ID: L1606127201
Report Date/Time: Wednesday, June 29, 2016 16:48:39
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**87.3928**	30.523	34.9	mg/L	0	Standard
Mg	24	56818.0	2.0	**66.6625**	1.829	2.7	mg/L	37	Standard
K	39	118.3	28.8	**0.1220**	0.059	48.6	mg/L	38	Standard
Ca	43	173.3	14.2	**7.4463**	1.537	20.6	mg/L	10	Standard
Fe	54	200.1	8.9	**0.0164**	0.019	116.3	mg/L	179	Standard
Fe	57	341.7	12.8	**-0.0156**	0.156	997.8	mg/L	335	Standard
> **Sc-1**	45	67137.0	4.7				mg/L	61029	Standard
Cl	35	31870.8	2.6				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	13232.4	3.4				ug/L	1547	Standard
P	31	22137.8	2.1				ug/L	15286	Standard
S	34	2160.2	8.0				ug/L	1572	Standard
Sr	88	146.7	24.2				ug/L	73	Standard
C	12	170.0	10.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	64.3	100.4				mg/L	6	Standard
Ho-1	165	55.0	78.7				mg/L	10	Standard
Er	166	50.0	60.0				mg/L	13	Standard
I	127	62898.8	4.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		126.291	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.237	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127201
Report Date/Time: Wednesday, June 29, 2016 16:48:39
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.859
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.945
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Sample ID: L1606127201
Report Date/Time: Wednesday, June 29, 2016 16:48:39
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127202
Sample Date/Time: Wednesday, June 29, 2016 16:49:33
Number of Replicates: 3
Autosampler Position: 222
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98788.0	4.1				ug/L	74699	Standard
	Be	9	13.3	57.3	**0.0192**	0.007	36.1	ug/L	8	Standard
	Al	27	43121483.9	2.2	**221.9923**	5.444	2.5	ug/L	1493	Standard
	Sc	45	66930.9	3.1				ug/L	61029	Standard
	Ti	47	1039.7	26.4	**2.4141**	0.661	27.4	ug/L	35	Standard
	V	51	49957.2	0.9	**5.8910**	0.067	1.1	ug/L	1435	Standard
	Cr	52	18344.1	1.8	**1.2538**	0.038	3.0	ug/L	7635	Standard
	Cr	53	24960.6	4.5	**24.5202**	1.080	4.4	ug/L	363	Standard
	Mn	55	17873.9	4.1	**2.2143**	0.103	4.7	ug/L	747	Standard
	Co	59	1405.4	6.7	**0.1586**	0.013	8.4	ug/L	234	Standard
	Ni	60	3876.2	1.9	**1.7810**	0.032	1.8	ug/L	215	Standard
	Cu	65	1407.4	2.7	**0.5921**	0.020	3.4	ug/L	104	Standard
	Zn	66	4623.4	0.6	**3.4401**	0.014	0.4	ug/L	363	Standard
>	**Ge**	72	538103.3	0.3				ug/L	495142	Standard
	As	75	2665.5	0.4	**2.2493**	0.008	0.3	ug/L	-50	Standard
	Se	82	859.9	5.4	**7.0717**	0.401	5.7	ug/L	14	Standard
	Se-1	77	1900.1	5.6	**23.3152**	1.302	5.6	ug/L	42	Standard
>	**Ga**	71	205.0	14.8				mg/L	42	Standard
	Rb	85	12111.5	1.9				ug/L	17	Standard
	Y	89	447538.5	3.5				ug/L	417098	Standard
>	**Rh**	103	186.7	5.6				ug/L	13	Standard
	Mo	98	4087.3	1.0	**1.4942**	0.026	1.7	ug/L	11	Standard
	Ag	107	258.7	10.6	**0.0259**	0.004	14.5	ug/L	68	Standard
	Cd	111	70.2	15.3	**0.0250**	0.005	21.6	mg/L	2	Standard
	Cd	114	224.5	9.2	**0.0407**	0.004	8.7	ug/L	12	Standard
>	**In**	115	600654.0	1.1				ug/L	587441	Standard
	Sn	118	765.0	14.6	**0.0335**	0.020	60.1	ug/L	497	Standard
	Sb	123	166.1	4.9	**0.0306**	0.002	5.1	ug/L	78	Standard
	Ba	135	52090.1	1.0	**24.8744**	0.038	0.2	ug/L	24	Standard
	Ce	140	4474.0	3.6				ug/L	25	Standard
>	**Tb**	159	974974.5	0.9				ug/L	939701	Standard
	Ho	165	80.0	27.2				ug/L	10	Standard
	Tl	203	82.0	10.0	**0.0082**	0.001	13.1	ug/L	4	Standard
	Tl	205	63.3	43.5	**0.0090**	0.004	42.2	ug/L	2	Standard
	Pb	206	1123.7	3.1	**0.1383**	0.006	4.7	ug/L	240	Standard
	Pb	207	914.0	2.3	**0.1218**	0.004	3.1	ug/L	198	Standard
	Pb	208	3779.9	2.5	**0.1267**	0.006	5.0	ug/L	883	Standard
	U	238	23648.8	1.1	**3.4170**	0.057	1.7	ug/L	3	Standard
>	**Bi**	209	526202.0	0.7				ug/L	523984	Standard

Sample ID: L1606127202
Report Date/Time: Wednesday, June 29, 2016 16:51:50
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**53.6294**	53.947	100.6	mg/L	0	Standard
Mg	24	113104.0	1.5	**133.1465**	4.642	3.5	mg/L	37	Standard
K	39	478.3	11.2	**0.8485**	0.082	9.7	mg/L	38	Standard
Ca	43	695.0	6.9	**37.5535**	3.460	9.2	mg/L	10	Standard
Fe	54	250.5	4.6	**0.0553**	0.004	6.4	mg/L	179	Standard
Fe	57	631.7	3.6	**0.8833**	0.015	1.7	mg/L	335	Standard
> **Sc-1**	45	66930.9	3.1				mg/L	61029	Standard
Cl	35	32927.1	2.7				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	41968.1	3.2				ug/L	1547	Standard
P	31	23413.1	2.8				ug/L	15286	Standard
S	34	2041.8	4.5				ug/L	1572	Standard
Sr	88	750.0	13.9				ug/L	73	Standard
C	12	126.7	19.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	103.6	24.2				mg/L	6	Standard
Ho-1	165	80.0	27.2				mg/L	10	Standard
Er	166	63.3	32.9				mg/L	13	Standard
I	127	36619.2	2.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.248	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.676	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127202
Report Date/Time: Wednesday, June 29, 2016 16:51:50
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.249
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.423
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606127202
Report Date/Time: Wednesday, June 29, 2016 16:51:50
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127901
Sample Date/Time: Wednesday, June 29, 2016 16:52:45
Number of Replicates: 3
Autosampler Position: 223
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94669.8	0.8				ug/L	74699	Standard
	Be	9	153.3	14.7	**0.1451**	0.021	14.4	ug/L	8	Standard
	Al	27	36828943.8	1.6	**197.7220**	2.539	1.3	ug/L	1493	Standard
	Sc	45	72494.8	2.9				ug/L	61029	Standard
	Ti	47	9823.2	1.5	**22.6070**	0.583	2.6	ug/L	35	Standard
	V	51	40160.9	1.2	**4.4642**	0.016	0.4	ug/L	1435	Standard
	Cr	52	82051.3	1.3	**8.6794**	0.190	2.2	ug/L	7635	Standard
	Cr	53	10588.7	4.0	**9.7108**	0.425	4.4	ug/L	363	Standard
	Mn	55	3156665.4	1.9	**397.2480**	11.605	2.9	ug/L	747	Standard
	Co	59	13202.1	1.7	**1.7150**	0.048	2.8	ug/L	234	Standard
	Ni	60	14584.3	1.0	**6.7029**	0.116	1.7	ug/L	215	Standard
	Cu	65	6796.9	2.1	**3.0972**	0.099	3.2	ug/L	104	Standard
	Zn	66	29601.9	0.7	**23.2962**	0.407	1.7	ug/L	363	Standard
>	**Ge**	72	565029.1	1.1				ug/L	495142	Standard
	As	75	1169.4	2.0	**0.9746**	0.024	2.5	ug/L	-50	Standard
	Se	82	54.4	2.8	**0.2964**	0.014	4.6	ug/L	14	Standard
	Se-1	77	95.3	8.8	**0.5209**	0.088	16.8	ug/L	42	Standard
>	**Ga**	71	2745.3	10.7				mg/L	42	Standard
	Rb	85	40779.8	2.2				ug/L	17	Standard
	Y	89	501434.2	0.6				ug/L	417098	Standard
>	**Rh**	103	116.7	40.7				ug/L	13	Standard
	Mo	98	6510.4	0.7	**2.1843**	0.030	1.4	ug/L	11	Standard
	Ag	107	146.0	1.2	**0.0072**	0.000	3.6	ug/L	68	Standard
	Cd	111	54.1	19.4	**0.0153**	0.004	28.2	mg/L	2	Standard
	Cd	114	169.8	20.2	**0.0276**	0.006	22.4	ug/L	12	Standard
>	**In**	115	655252.5	1.2				ug/L	587441	Standard
	Sn	118	1435.1	4.4	**0.1255**	0.007	5.6	ug/L	497	Standard
	Sb	123	419.8	7.1	**0.0764**	0.005	6.8	ug/L	78	Standard
	Ba	135	167537.6	1.4	**73.3690**	0.619	0.8	ug/L	24	Standard
	Ce	140	144147.5	0.8				ug/L	25	Standard
>	**Tb**	159	1041877.9	1.3				ug/L	939701	Standard
	Ho	165	2731.9	2.9				ug/L	10	Standard
	Tl	203	237.3	5.5	**0.0244**	0.001	5.3	ug/L	4	Standard
	Tl	205	228.3	8.3	**0.0292**	0.002	8.4	ug/L	2	Standard
	Pb	206	7357.1	0.9	**1.1818**	0.018	1.5	ug/L	240	Standard
	Pb	207	5920.5	2.7	**1.0648**	0.026	2.5	ug/L	198	Standard
	Pb	208	25273.6	1.1	**1.1521**	0.016	1.4	ug/L	883	Standard
	U	238	16614.4	0.8	**2.1915**	0.022	1.0	ug/L	3	Standard
>	**Bi**	209	575969.2	0.6				ug/L	523984	Standard

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:55:02
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**32.3968**	28.087	86.7	mg/L	0	Standard
Mg	24	3568.8	4.7	**3.8119**	0.091	2.4	mg/L	37	Standard
K	39	285.0	24.6	**0.4167**	0.132	31.6	mg/L	38	Standard
Ca	43	908.4	6.1	**45.7783**	2.516	5.5	mg/L	10	Standard
Fe	54	3605.6	1.3	**2.4364**	0.072	2.9	mg/L	179	Standard
Fe	57	1446.7	0.5	**3.0674**	0.109	3.5	mg/L	335	Standard
Sc-1	45	72494.8	2.9				mg/L	61029	Standard
Cl	35	31293.6	1.2				ug/L	20195	Standard
Kr	83	2.3	49.5				ug/L	4	Standard
Br	81	8015.5	4.3				ug/L	1547	Standard
P	31	26808.7	3.1				ug/L	15286	Standard
S	34	2353.5	0.9				ug/L	1572	Standard
Sr	88	506.7	5.4				ug/L	73	Standard
C	12	183.3	20.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	3	Standard
Dy	164	4522.3	4.2				mg/L	6	Standard
Ho-1	165	2731.9	2.9				mg/L	10	Standard
Er	166	2333.5	8.8				mg/L	13	Standard
I	127	38263.4	7.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		126.735	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.114	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:55:02
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	111.544
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	109.921
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:55:02
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127901
Sample Date/Time: Wednesday, June 29, 2016 16:55:56
Number of Replicates: 3
Autosampler Position: 224
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94757.1	0.6				ug/L	74699	Standard
	Be	9	16.7	34.6	**0.0226**	0.005	23.2	ug/L	8	Standard
	Al	27	809583.1	0.4	**4.3294**	0.010	0.2	ug/L	1493	Standard
	Sc	45	70063.3	2.2				ug/L	61029	Standard
	Ti	47	239.7	2.8	**0.4418**	0.016	3.7	ug/L	35	Standard
	V	51	2635.2	3.3	**0.1083**	0.010	9.6	ug/L	1435	Standard
	Cr	52	10594.0	0.1	**0.2214**	0.004	1.7	ug/L	7635	Standard
	Cr	53	1061.7	10.6	**0.6708**	0.107	16.0	ug/L	363	Standard
	Mn	55	69843.0	1.9	**8.5894**	0.152	1.8	ug/L	747	Standard
	Co	59	472.7	1.3	**0.0251**	0.001	3.3	ug/L	234	Standard
	Ni	60	497.3	0.6	**0.1075**	0.002	1.4	ug/L	215	Standard
	Cu	65	314.7	2.4	**0.0436**	0.003	7.6	ug/L	104	Standard
	Zn	66	3346.7	3.2	**2.2142**	0.089	4.0	ug/L	363	Standard
>	**Ge**	72	568460.2	0.3				ug/L	495142	Standard
	As	75	-20.7	45.1	**0.0435**	0.007	16.6	ug/L	-50	Standard
	Se	82	18.3	17.6	**0.0072**	0.025	350.3	ug/L	14	Standard
	Se-1	77	58.0	6.2	**0.0694**	0.043	62.2	ug/L	42	Standard
>	**Ga**	71	123.3	16.9				mg/L	42	Standard
	Rb	85	875.0	6.6				ug/L	17	Standard
	Y	89	490331.7	1.2				ug/L	417098	Standard
>	**Rh**	103	16.7	45.8				ug/L	13	Standard
	Mo	98	175.2	3.1	**0.0513**	0.001	2.5	ug/L	11	Standard
	Ag	107	187.0	5.8	**0.0120**	0.001	10.3	ug/L	68	Standard
	Cd	111	8.4	36.2	**-0.0044**	0.001	27.6	mg/L	2	Standard
	Cd	114	20.0	63.9	**0.0013**	0.002	167.4	ug/L	12	Standard
>	**In**	115	676596.2	0.9				ug/L	587441	Standard
	Sn	118	621.7	13.0	**-0.0023**	0.012	549.8	ug/L	497	Standard
	Sb	123	148.0	17.5	**0.0233**	0.005	19.5	ug/L	78	Standard
	Ba	135	3597.4	2.3	**1.5099**	0.025	1.6	ug/L	24	Standard
	Ce	140	3133.7	5.7				ug/L	25	Standard
>	**Tb**	159	1053037.0	0.7				ug/L	939701	Standard
	Ho	165	73.3	37.6				ug/L	10	Standard
	Tl	203	38.7	9.8	**0.0024**	0.000	14.0	ug/L	4	Standard
	Tl	205	30.0	57.7	**0.0039**	0.002	55.8	ug/L	2	Standard
	Pb	206	500.3	0.4	**0.0113**	0.001	11.7	ug/L	240	Standard
	Pb	207	394.7	10.5	**0.0033**	0.008	256.3	ug/L	198	Standard
	Pb	208	1728.4	3.9	**0.0071**	0.003	37.0	ug/L	883	Standard
	U	238	362.3	7.5	**0.0420**	0.004	8.9	ug/L	3	Standard
>	**Bi**	209	594905.3	1.7				ug/L	523984	Standard

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:58:13
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.1214**	29.647	173.2	mg/L	0	Standard
Mg	24	156.7	18.1	**0.1107**	0.028	25.7	mg/L	37	Standard
K	39	30.0	60.1	**-0.0567**	0.034	59.1	mg/L	38	Standard
Ca	43	53.3	5.4	**0.4139**	0.213	51.5	mg/L	10	Standard
Fe	54	254.1	20.9	**0.0493**	0.038	78.1	mg/L	179	Standard
Fe	57	286.7	7.0	**-0.2249**	0.071	31.4	mg/L	335	Standard
> **Sc-1**	45	70063.3	2.2				mg/L	61029	Standard
Cl	35	33357.3	0.7				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	2703.6	11.1				ug/L	1547	Standard
P	31	21573.7	6.4				ug/L	15286	Standard
S	34	2518.5	1.8				ug/L	1572	Standard
Sr	88	90.0	20.0				ug/L	73	Standard
C	12	76.7	7.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	206.2	21.0				mg/L	6	Standard
Ho-1	165	73.3	37.6				mg/L	10	Standard
Er	166	80.0	12.5				mg/L	13	Standard
I	127	6318.0	20.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		126.852	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.807	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:58:13
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	115.177
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	113.535
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 16:59:09
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92097.4	0.2				ug/L	74699	Standard
	Be	9	59332.7	2.3	**54.6708**	1.340	2.5	ug/L	8	Standard
	Al	27	8920018.4	1.0	**49.2175**	0.589	1.2	ug/L	1493	Standard
	Sc	45	68374.0	2.7				ug/L	61029	Standard
	Ti	47	44746.6	0.7	**103.2626**	1.432	1.4	ug/L	35	Standard
	V	51	467593.6	0.9	**54.0107**	0.396	0.7	ug/L	1435	Standard
	Cr	52	444978.1	1.4	**51.5506**	1.111	2.2	ug/L	7635	Standard
	Cr	53	54491.2	2.0	**51.2973**	1.411	2.7	ug/L	363	Standard
	Mn	55	420710.6	2.8	**52.7641**	1.799	3.4	ug/L	747	Standard
	Co	59	425350.8	2.7	**56.3649**	1.831	3.2	ug/L	234	Standard
	Ni	60	109363.3	1.2	**51.0215**	0.971	1.9	ug/L	215	Standard
	Cu	65	108714.3	1.3	**51.0413**	1.018	2.0	ug/L	104	Standard
	Zn	66	63872.0	1.2	**50.7455**	1.020	2.0	ug/L	363	Standard
>	**Ge**	72	565538.7	0.9				ug/L	495142	Standard
	As	75	65590.6	1.0	**51.3298**	0.623	1.2	ug/L	-50	Standard
	Se	82	6676.5	1.4	**53.1245**	0.865	1.6	ug/L	14	Standard
	Se-1	77	4570.7	2.8	**54.1793**	1.989	3.7	ug/L	42	Standard
>	**Ga**	71	90.0	25.5				mg/L	42	Standard
	Rb	85	980.0	1.8				ug/L	17	Standard
	Y	89	481101.3	1.3				ug/L	417098	Standard
>	**Rh**	103	36.7	20.8				ug/L	13	Standard
	Mo	98	317636.1	2.3	**106.0178**	2.920	2.8	ug/L	11	Standard
	Ag	107	377887.3	1.0	**51.5689**	0.593	1.2	ug/L	68	Standard
	Cd	111	119799.3	1.9	**50.9674**	1.189	2.3	mg/L	2	Standard
	Cd	114	291345.5	1.8	**50.5384**	1.150	2.3	ug/L	12	Standard
>	**In**	115	660345.6	0.5				ug/L	587441	Standard
	Sn	118	333887.3	1.0	**50.7546**	0.756	1.5	ug/L	497	Standard
	Sb	123	263770.3	1.5	**50.2579**	0.987	2.0	ug/L	78	Standard
	Ba	135	110751.0	1.4	**48.1238**	0.926	1.9	ug/L	24	Standard
	Ce	140	100.0	13.2				ug/L	25	Standard
>	**Tb**	159	1027328.3	1.0				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	469667.3	0.3	**51.2282**	0.784	1.5	ug/L	4	Standard
	Tl	205	432915.4	1.7	**54.4135**	1.621	3.0	ug/L	2	Standard
	Pb	206	304012.7	1.3	**51.1045**	1.311	2.6	ug/L	240	Standard
	Pb	207	269196.3	1.3	**50.9328**	1.301	2.6	ug/L	198	Standard
	Pb	208	1105976.1	1.5	**52.9652**	1.506	2.8	ug/L	883	Standard
	U	238	408284.4	0.8	**53.3305**	1.093	2.0	ug/L	3	Standard
>	**Bi**	209	582837.4	1.3				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:01:26
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4664.1	3.2	**5.3139**	0.299	5.6	mg/L	37	Standard
K	39	2066.8	2.7	**3.9658**	0.176	4.4	mg/L	38	Standard
Ca	43	83.3	3.5	**2.1752**	0.111	5.1	mg/L	10	Standard
Fe	54	7259.4	3.7	**5.3547**	0.102	1.9	mg/L	179	Standard
Fe	57	2016.8	2.5	**5.0496**	0.304	6.0	mg/L	335	Standard
Sc-1	45	68374.0	2.7				mg/L	61029	Standard
Cl	35	31290.3	2.5				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	2203.5	6.4				ug/L	1547	Standard
P	31	22518.4	0.7				ug/L	15286	Standard
S	34	2795.3	6.4				ug/L	1572	Standard
Sr	88	58.3	13.1				ug/L	73	Standard
C	12	70.0	37.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	25.1	45.0				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	33.3	34.6				mg/L	13	Standard
I	127	4404.0	17.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	109.342		
Al	27	98.435		
Sc	45			
Ti	47	103.263		
V	51	108.021		
Cr	52	103.101		
Cr	53			
Mn	55	105.528		
Co	59	112.730		
Ni	60	102.043		
Cu	65	102.083		
Zn	66	101.491		
Ge	72		114.217	
As	75	102.660		
Se	82	106.249		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:01:26
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	106.018		
Ag	107	103.138		
Cd	111	101.935		
Cd	114			
In	115		112.411	
Sn	118	101.509		
Sb	123	100.516		
Ba	135	96.248		
Ce	140			
Tb	159			
Ho	165			
Tl	203	102.456		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.930		
U	238	106.661		
Bi	209		111.232	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:01:26
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 17:02:20
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92622.3	1.7				ug/L	74699	Standard
	Be	9	38.3	64.3	**0.0426**	0.023	52.8	ug/L	8	Standard
	Al	27	23370.4	151.2	**0.1159**	0.196	168.9	ug/L	1493	Standard
	Sc	45	68951.6	1.3				ug/L	61029	Standard
	Ti	47	40.3	21.1	**-0.0139**	0.019	138.6	ug/L	35	Standard
	V	51	1665.7	22.9	**0.0015**	0.044	2840.4	ug/L	1435	Standard
	Cr	52	7712.6	2.1	**-0.0923**	0.036	39.5	ug/L	7635	Standard
	Cr	53	583.3	10.9	**0.2339**	0.049	21.0	ug/L	363	Standard
	Mn	55	4624.3	138.3	**0.4433**	0.816	184.2	ug/L	747	Standard
	Co	59	398.3	65.9	**0.0165**	0.035	212.9	ug/L	234	Standard
	Ni	60	246.3	49.7	**-0.0062**	0.058	932.9	ug/L	215	Standard
	Cu	65	154.7	27.9	**-0.0294**	0.020	68.5	ug/L	104	Standard
	Zn	66	470.3	19.7	**-0.0693**	0.075	107.9	ug/L	363	Standard
>	**Ge**	72	553538.3	2.4				ug/L	495142	Standard
	As	75	-38.0	63.4	**0.0295**	0.019	62.8	ug/L	-50	Standard
	Se	82	20.6	12.7	**0.0307**	0.026	83.7	ug/L	14	Standard
	Se-1	77	63.0	12.0	**0.1490**	0.085	57.3	ug/L	42	Standard
>	**Ga**	71	71.7	14.5				mg/L	42	Standard
	Rb	85	25.0	0.0				ug/L	17	Standard
	Y	89	484828.2	0.4				ug/L	417098	Standard
>	**Rh**	103	11.7	65.5				ug/L	13	Standard
	Mo	98	326.7	26.1	**0.0989**	0.027	26.9	ug/L	11	Standard
	Ag	107	215.0	76.0	**0.0152**	0.021	139.6	ug/L	68	Standard
	Cd	111	16.1	49.4	**-0.0013**	0.003	242.1	mg/L	2	Standard
	Cd	114	45.0	78.6	**0.0054**	0.006	109.2	ug/L	12	Standard
>	**In**	115	687277.7	0.8				ug/L	587441	Standard
	Sn	118	931.7	18.6	**0.0418**	0.027	63.5	ug/L	497	Standard
	Sb	123	489.1	40.1	**0.0853**	0.036	41.8	ug/L	78	Standard
	Ba	135	52.0	47.0	**0.0057**	0.010	177.6	ug/L	24	Standard
	Ce	140	81.7	78.0				ug/L	25	Standard
>	**Tb**	159	1059717.9	1.6				ug/L	939701	Standard
	Ho	165	6.7	86.6				ug/L	10	Standard
	Tl	203	97.3	80.9	**0.0086**	0.008	97.3	ug/L	4	Standard
	Tl	205	126.7	77.5	**0.0158**	0.012	76.4	ug/L	2	Standard
	Pb	206	434.7	47.9	**0.0006**	0.034	5998.1	ug/L	240	Standard
	Pb	207	327.7	41.4	**-0.0091**	0.025	275.6	ug/L	198	Standard
	Pb	208	1259.4	26.8	**-0.0148**	0.016	107.4	ug/L	883	Standard
	U	238	94.7	104.8	**0.0077**	0.013	165.1	ug/L	3	Standard
>	**Bi**	209	593862.7	2.1				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:04:37
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	88.3	93.4	**0.0352**	0.093	262.9	mg/L	37	Standard
K	39	41.7	30.2	**-0.0325**	0.025	77.1	mg/L	38	Standard
Ca	43	21.7	58.1	**-1.3088**	0.708	54.1	mg/L	10	Standard
Fe	54	169.3	10.4	**-0.0113**	0.012	105.6	mg/L	179	Standard
Fe	57	255.0	7.1	**-0.3067**	0.061	20.0	mg/L	335	Standard
> **Sc-1**	45	68951.6	1.3				mg/L	61029	Standard
Cl	35	30862.1	1.0				ug/L	20195	Standard
Kr	83	3.7	31.5				ug/L	4	Standard
Br	81	2130.2	6.5				ug/L	1547	Standard
P	31	22533.4	1.3				ug/L	15286	Standard
S	34	2576.9	2.2				ug/L	1572	Standard
Sr	88	61.7	20.4				ug/L	73	Standard
C	12	70.0	62.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.1	95.9				mg/L	6	Standard
Ho-1	165	6.7	86.6				mg/L	10	Standard
Er	166	33.3	69.3				mg/L	13	Standard
I	127	3348.7	10.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.794	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:04:37
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.995
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.336
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Mn	55	

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:04:37
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127902 WG574001-03
Sample Date/Time: Wednesday, June 29, 2016 17:05:33
Number of Replicates: 3
Autosampler Position: 225
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	97844.9	1.2				ug/L	74699	Standard
	Be	9	628.3	5.4	**0.5525**	0.030	5.3	ug/L	8	Standard
	Al	27	67592650.7	0.8	**351.1766**	6.777	1.9	ug/L	1493	Standard
	Sc	45	76196.2	3.2				ug/L	61029	Standard
	Ti	47	35031.5	1.8	**85.4603**	1.454	1.7	ug/L	35	Standard
	V	51	127107.8	0.4	**15.3856**	0.087	0.6	ug/L	1435	Standard
	Cr	52	144594.0	0.9	**17.0395**	0.185	1.1	ug/L	7635	Standard
	Cr	53	17708.7	3.4	**17.4084**	0.576	3.3	ug/L	363	Standard
	Mn	55	9638149.6	0.5	**1281.4698**	6.113	0.5	ug/L	747	Standard
	Co	59	44221.4	1.1	**6.1626**	0.079	1.3	ug/L	234	Standard
	Ni	60	29440.2	0.7	**14.4336**	0.102	0.7	ug/L	215	Standard
	Cu	65	18715.2	1.1	**9.2058**	0.118	1.3	ug/L	104	Standard
	Zn	66	49911.7	1.0	**41.8483**	0.443	1.1	ug/L	363	Standard
>	**Ge**	72	534821.8	0.2				ug/L	495142	Standard
	As	75	5363.3	0.7	**4.4926**	0.039	0.9	ug/L	-50	Standard
	Se	82	81.7	4.2	**0.5509**	0.030	5.5	ug/L	14	Standard
	Se-1	77	90.0	5.1	**0.5184**	0.056	10.8	ug/L	42	Standard
>	**Ga**	71	10246.8	1.6				mg/L	42	Standard
	Rb	85	128176.6	1.2				ug/L	17	Standard
	Y	89	611826.2	1.2				ug/L	417098	Standard
>	**Rh**	103	203.3	5.1				ug/L	13	Standard
	Mo	98	8459.2	1.1	**2.8837**	0.063	2.2	ug/L	11	Standard
	Ag	107	335.7	32.7	**0.0339**	0.015	43.7	ug/L	68	Standard
	Cd	111	167.4	13.2	**0.0649**	0.009	13.6	mg/L	2	Standard
	Cd	114	471.4	29.4	**0.0814**	0.024	29.1	ug/L	12	Standard
>	**In**	115	645352.1	1.1				ug/L	587441	Standard
	Sn	118	2433.5	1.0	**0.2845**	0.006	2.0	ug/L	497	Standard
	Sb	123	747.2	5.1	**0.1415**	0.006	4.4	ug/L	78	Standard
	Ba	135	186708.8	1.1	**83.0208**	0.490	0.6	ug/L	24	Standard
	Ce	140	555105.9	0.4				ug/L	25	Standard
>	**Tb**	159	1017797.9	1.6				ug/L	939701	Standard
	Ho	165	10682.1	1.8				ug/L	10	Standard
	Tl	203	801.4	12.7	**0.0906**	0.011	12.4	ug/L	4	Standard
	Tl	205	726.7	5.1	**0.0967**	0.005	5.6	ug/L	2	Standard
	Pb	206	28687.5	0.5	**5.0315**	0.023	0.5	ug/L	240	Standard
	Pb	207	23046.8	1.0	**4.5438**	0.005	0.1	ug/L	198	Standard
	Pb	208	100519.8	0.6	**5.0199**	0.074	1.5	ug/L	883	Standard
	U	238	32058.9	0.6	**4.4211**	0.050	1.1	ug/L	3	Standard
>	**Bi**	209	551473.5	0.9				ug/L	523984	Standard

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:07:50
Page 1

Approved: July 04, 2016

Na	23	16.7	17.3	**153.4719**	22.118	14.4	mg/L	0	Standard
Mg	24	8050.5	4.4	**8.2571**	0.176	2.1	mg/L	37	Standard
K	39	990.0	1.3	**1.6393**	0.041	2.5	mg/L	38	Standard
Ca	43	1383.4	8.1	**67.3924**	3.470	5.1	mg/L	10	Standard
Fe	54	12019.5	0.9	**8.0301**	0.303	3.8	mg/L	179	Standard
Fe	57	3888.8	1.5	**9.5221**	0.470	4.9	mg/L	335	Standard
Sc-1	45	76196.2	3.2				mg/L	61029	Standard
Cl	35	38282.3	2.8				ug/L	20195	Standard
Kr	83	1.7	124.9				ug/L	4	Standard
Br	81	9239.5	4.1				ug/L	1547	Standard
P	31	33428.5	3.8				ug/L	15286	Standard
S	34	2046.8	10.4				ug/L	1572	Standard
Sr	88	885.0	8.4				ug/L	73	Standard
C	12	223.3	22.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	3	Standard
Dy	164	18454.9	2.8				mg/L	6	Standard
Ho-1	165	10682.1	1.8				mg/L	10	Standard
Er	166	9956.6	4.7				mg/L	13	Standard
I	127	65381.5	9.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		130.986	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		108.014	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:07:50
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.858
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.246
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:07:50
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127902 WG574001-03
Sample Date/Time: Wednesday, June 29, 2016 17:08:45
Number of Replicates: 3
Autosampler Position: 226
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	97333.3	1.1				ug/L	74699	Standard
	Be	9	23.3	32.7	**0.0279**	0.007	23.3	ug/L	8	Standard
	Al	27	1471688.9	2.0	**7.6717**	0.099	1.3	ug/L	1493	Standard
	Sc	45	68584.9	0.8				ug/L	61029	Standard
	Ti	47	736.0	3.1	**1.5714**	0.076	4.9	ug/L	35	Standard
	V	51	4360.2	2.0	**0.3040**	0.018	5.9	ug/L	1435	Standard
	Cr	52	10891.2	1.4	**0.2475**	0.011	4.5	ug/L	7635	Standard
	Cr	53	901.7	3.1	**0.5139**	0.018	3.4	ug/L	363	Standard
	Mn	55	209342.7	1.0	**25.8667**	0.657	2.5	ug/L	747	Standard
	Co	59	1184.0	8.5	**0.1180**	0.016	13.3	ug/L	234	Standard
	Ni	60	823.4	2.4	**0.2566**	0.015	5.7	ug/L	215	Standard
	Cu	65	604.0	5.8	**0.1771**	0.017	9.7	ug/L	104	Standard
	Zn	66	3536.4	2.1	**2.3459**	0.038	1.6	ug/L	363	Standard
>	**Ge**	72	572433.6	1.6				ug/L	495142	Standard
	As	75	56.5	19.8	**0.1032**	0.008	8.2	ug/L	-50	Standard
	Se	82	22.7	8.9	**0.0409**	0.015	36.5	ug/L	14	Standard
	Se-1	77	56.3	3.7	**0.0451**	0.032	71.6	ug/L	42	Standard
>	**Ga**	71	286.7	20.8				mg/L	42	Standard
	Rb	85	2816.9	8.1				ug/L	17	Standard
	Y	89	483993.7	0.1				ug/L	417098	Standard
>	**Rh**	103	11.7	65.5				ug/L	13	Standard
	Mo	98	255.8	9.6	**0.0784**	0.008	10.6	ug/L	11	Standard
	Ag	107	215.7	39.7	**0.0161**	0.012	72.2	ug/L	68	Standard
	Cd	111	33.2	125.7	**0.0061**	0.018	290.4	mg/L	2	Standard
	Cd	114	83.8	140.0	**0.0123**	0.020	164.2	ug/L	12	Standard
>	**In**	115	670122.5	0.4				ug/L	587441	Standard
	Sn	118	730.0	9.6	**0.0149**	0.011	73.4	ug/L	497	Standard
	Sb	123	198.5	22.7	**0.0331**	0.009	26.1	ug/L	78	Standard
	Ba	135	4104.6	4.5	**1.7422**	0.086	5.0	ug/L	24	Standard
	Ce	140	11919.7	3.4				ug/L	25	Standard
>	**Tb**	159	1040073.5	0.3				ug/L	939701	Standard
	Ho	165	248.3	25.2				ug/L	10	Standard
	Tl	203	119.3	122.5	**0.0111**	0.016	142.6	ug/L	4	Standard
	Tl	205	45.0	44.4	**0.0058**	0.003	43.8	ug/L	2	Standard
	Pb	206	1027.4	11.1	**0.0985**	0.020	20.7	ug/L	240	Standard
	Pb	207	828.0	1.8	**0.0839**	0.004	4.9	ug/L	198	Standard
	Pb	208	3389.1	2.4	**0.0853**	0.005	6.3	ug/L	883	Standard
	U	238	690.0	3.8	**0.0840**	0.004	4.3	ug/L	3	Standard
>	**Bi**	209	594228.5	1.0				ug/L	523984	Standard

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:11:02
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	215.0	6.2	**0.1819**	0.016	8.8	mg/L	37	Standard
K	39	53.3	14.3	**-0.0093**	0.014	153.2	mg/L	38	Standard
Ca	43	48.3	15.8	**0.1914**	0.411	214.9	mg/L	10	Standard
Fe	54	443.5	15.8	**0.1966**	0.055	28.1	mg/L	179	Standard
Fe	57	323.3	11.7	**-0.0960**	0.118	122.9	mg/L	335	Standard
> **Sc-1**	45	68584.9	0.8				mg/L	61029	Standard
Cl	35	29711.8	0.5				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	2166.8	7.4				ug/L	1547	Standard
P	31	19419.1	3.8				ug/L	15286	Standard
S	34	2206.8	3.0				ug/L	1572	Standard
Sr	88	96.7	19.6				ug/L	73	Standard
C	12	46.7	49.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	390.3	4.3				mg/L	6	Standard
Ho-1	165	248.3	25.2				mg/L	10	Standard
Er	166	203.3	36.3				mg/L	13	Standard
I	127	5617.8	25.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		130.301	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.610	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:11:02
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.075
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.406
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127903S WG574001-09
Sample Date/Time: Wednesday, June 29, 2016 17:11:56
Number of Replicates: 3
Autosampler Position: 227
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	103734.0	1.8				ug/L	74699	Standard
	Be	9	62659.4	2.5	**51.2580**	0.858	1.7	ug/L	8	Standard
	Al	27	65559814.6	2.9	**321.3254**	12.360	3.8	ug/L	1493	Standard
	Sc	45	78765.9	1.8				ug/L	61029	Standard
	Ti	47	33652.0	1.4	**81.0287**	2.211	2.7	ug/L	35	Standard
	V	51	549023.7	0.9	**66.2300**	1.368	2.1	ug/L	1435	Standard
	Cr	52	510507.6	0.8	**61.9303**	1.817	2.9	ug/L	7635	Standard
	Cr	53	62490.3	1.3	**61.4602**	1.770	2.9	ug/L	363	Standard
	Mn	55	9311676.0	0.4	**1221.9970**	25.998	2.1	ug/L	747	Standard
	Co	59	420012.9	0.8	**58.0870**	1.492	2.6	ug/L	234	Standard
	Ni	60	121605.8	1.9	**59.2404**	2.250	3.8	ug/L	215	Standard
	Cu	65	110409.9	1.5	**54.1128**	1.822	3.4	ug/L	104	Standard
	Zn	66	101498.2	0.9	**84.4641**	2.491	2.9	ug/L	363	Standard
>	**Ge**	72	542031.1	2.4				ug/L	495142	Standard
	As	75	66455.1	1.4	**54.2648**	0.548	1.0	ug/L	-50	Standard
	Se	82	6239.8	1.8	**51.8120**	1.305	2.5	ug/L	14	Standard
	Se-1	77	4136.2	2.8	**51.1231**	1.591	3.1	ug/L	42	Standard
>	**Ga**	71	10863.9	0.3				mg/L	42	Standard
	Rb	85	119392.1	2.4				ug/L	17	Standard
	Y	89	586391.1	1.2				ug/L	417098	Standard
>	**Rh**	103	175.0	15.1				ug/L	13	Standard
	Mo	98	5740.7	2.3	**1.9719**	0.058	3.0	ug/L	11	Standard
	Ag	107	333249.6	2.0	**46.9378**	1.203	2.6	ug/L	68	Standard
	Cd	111	109090.4	0.8	**47.8987**	0.737	1.5	mg/L	2	Standard
	Cd	114	263561.8	0.5	**47.1881**	0.999	2.1	ug/L	12	Standard
>	**In**	115	639905.2	1.9				ug/L	587441	Standard
	Sn	118	1925.1	5.8	**0.2078**	0.017	8.3	ug/L	497	Standard
	Sb	123	124661.9	1.5	**24.5159**	0.713	2.9	ug/L	78	Standard
	Ba	135	268706.4	1.4	**120.5245**	2.161	1.8	ug/L	24	Standard
	Ce	140	538323.1	0.7				ug/L	25	Standard
>	**Tb**	159	1045169.0	2.0				ug/L	939701	Standard
	Ho	165	11317.6	5.4				ug/L	10	Standard
	Tl	203	438870.4	1.4	**49.1992**	0.166	0.3	ug/L	4	Standard
	Tl	205	398898.5	1.5	**51.5243**	0.281	0.5	ug/L	2	Standard
	Pb	206	314887.1	0.5	**54.4073**	0.494	0.9	ug/L	240	Standard
	Pb	207	267442.5	0.8	**52.0088**	0.664	1.3	ug/L	198	Standard
	Pb	208	1101699.4	1.8	**54.2208**	0.536	1.0	ug/L	883	Standard
	U	238	415059.0	1.7	**55.7208**	0.743	1.3	ug/L	3	Standard
>	**Bi**	209	566992.1	1.1				ug/L	523984	Standard

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:14:13
Page 1

Approved: July 04, 2016

Na	23	8.3	69.3	**74.8950**	52.957	70.7	mg/L	0	Standard
Mg	24	7490.2	5.9	**7.4319**	0.549	7.4	mg/L	37	Standard
K	39	995.0	6.5	**1.5912**	0.134	8.4	mg/L	38	Standard
Ca	43	1293.4	1.8	**60.7979**	1.957	3.2	mg/L	10	Standard
Fe	54	10550.1	1.1	**6.7940**	0.164	2.4	mg/L	179	Standard
Fe	57	3448.7	1.2	**8.0106**	0.227	2.8	mg/L	335	Standard
> **Sc-1**	45	78765.9	1.8				mg/L	61029	Standard
Cl	35	36168.5	1.3				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	7995.5	8.9				ug/L	1547	Standard
P	31	34405.7	2.4				ug/L	15286	Standard
S	34	1880.1	7.8				ug/L	1572	Standard
Sr	88	881.7	7.6				ug/L	73	Standard
C	12	140.0	46.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	3	Standard
Dy	164	18849.9	4.6				mg/L	6	Standard
Ho-1	165	11317.6	5.4				mg/L	10	Standard
Er	166	9863.2	2.6				mg/L	13	Standard
I	127	61177.5	10.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		138.870	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		109.470	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:14:13
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	108.931
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.208
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:14:13
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127903S WG574001-09
Sample Date/Time: Wednesday, June 29, 2016 17:15:08
Number of Replicates: 3
Autosampler Position: 228
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100014.2	2.8				ug/L	74699	Standard
	Be	9	1316.7	11.4	**1.1231**	0.100	8.9	ug/L	8	Standard
	Al	27	1420391.6	2.6	**7.2057**	0.041	0.6	ug/L	1493	Standard
	Sc	45	65355.8	3.5				ug/L	61029	Standard
	Ti	47	751.7	2.2	**1.7435**	0.043	2.4	ug/L	35	Standard
	V	51	12553.2	2.5	**1.3568**	0.030	2.2	ug/L	1435	Standard
	Cr	52	18458.6	1.1	**1.3028**	0.035	2.7	ug/L	7635	Standard
	Cr	53	1771.8	3.3	**1.4619**	0.098	6.7	ug/L	363	Standard
	Mn	55	203580.1	1.5	**27.1635**	0.320	1.2	ug/L	747	Standard
	Co	59	9135.1	1.0	**1.2550**	0.021	1.7	ug/L	234	Standard
	Ni	60	2779.3	1.4	**1.2628**	0.031	2.5	ug/L	215	Standard
	Cu	65	2563.2	1.7	**1.1828**	0.027	2.3	ug/L	104	Standard
	Zn	66	4447.0	0.1	**3.3488**	0.082	2.5	ug/L	363	Standard
>	**Ge**	72	530183.0	2.3				ug/L	495142	Standard
	As	75	1256.2	4.8	**1.1077**	0.065	5.9	ug/L	-50	Standard
	Se	82	144.5	11.0	**1.0936**	0.160	14.6	ug/L	14	Standard
	Se-1	77	149.3	5.6	**1.2873**	0.101	7.8	ug/L	42	Standard
>	**Ga**	71	291.7	12.0				mg/L	42	Standard
	Rb	85	2473.5	8.7				ug/L	17	Standard
	Y	89	480063.9	1.1				ug/L	417098	Standard
>	**Rh**	103	18.3	15.7				ug/L	13	Standard
	Mo	98	168.4	5.0	**0.0496**	0.003	5.9	ug/L	11	Standard
	Ag	107	7406.8	1.4	**0.9815**	0.012	1.2	ug/L	68	Standard
	Cd	111	2324.9	2.0	**0.9649**	0.013	1.3	mg/L	2	Standard
	Cd	114	5499.5	2.5	**0.9362**	0.028	2.9	ug/L	12	Standard
>	**In**	115	671407.0	1.0				ug/L	587441	Standard
	Sn	118	800.0	6.0	**0.0251**	0.007	29.1	ug/L	497	Standard
	Sb	123	3310.0	4.5	**0.6160**	0.025	4.0	ug/L	78	Standard
	Ba	135	5738.4	0.8	**2.4371**	0.009	0.4	ug/L	24	Standard
	Ce	140	11352.6	1.4				ug/L	25	Standard
>	**Tb**	159	1030086.3	0.3				ug/L	939701	Standard
	Ho	165	295.0	7.8				ug/L	10	Standard
	Tl	203	9225.5	0.2	**0.9995**	0.013	1.3	ug/L	4	Standard
	Tl	205	8459.0	5.0	**1.0575**	0.041	3.9	ug/L	2	Standard
	Pb	206	6958.9	0.8	**1.0943**	0.025	2.3	ug/L	240	Standard
	Pb	207	5899.8	0.2	**1.0421**	0.014	1.4	ug/L	198	Standard
	Pb	208	25022.5	1.7	**1.1197**	0.025	2.3	ug/L	883	Standard
	U	238	8748.9	2.2	**1.1326**	0.025	2.2	ug/L	3	Standard
>	**Bi**	209	585775.3	1.4				ug/L	523984	Standard

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:17:25
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	185.0	5.4	**0.1579**	0.009	5.9	mg/L	37	Standard
K	39	46.7	32.7	**-0.0170**	0.035	202.9	mg/L	38	Standard
Ca	43	53.3	5.4	**0.6286**	0.271	43.1	mg/L	10	Standard
Fe	54	400.2	11.9	**0.1796**	0.048	26.7	mg/L	179	Standard
Fe	57	335.0	9.8	**-0.0119**	0.080	674.7	mg/L	335	Standard
Sc-1	45	65355.8	3.5				mg/L	61029	Standard
Cl	35	29320.3	1.0				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	2030.1	9.5				ug/L	1547	Standard
P	31	19806.3	2.6				ug/L	15286	Standard
S	34	2145.2	0.2				ug/L	1572	Standard
Sr	88	86.7	23.3				ug/L	73	Standard
C	12	80.0	66.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	379.6	21.3				mg/L	6	Standard
Ho-1	165	295.0	7.8				mg/L	10	Standard
Er	166	216.7	11.6				mg/L	13	Standard
I	127	5157.6	28.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		133.890	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		107.077	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:17:25
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.294
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.792
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127904SD WG574001-10
Sample Date/Time: Wednesday, June 29, 2016 17:18:19
Number of Replicates: 3
Autosampler Position: 229
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100168.4	0.8				ug/L	74699	Standard
	Be	9	61913.0	2.3	**52.4502**	1.133	2.2	ug/L	8	Standard
	Al	27	66193036.9	1.4	**335.9039**	6.576	2.0	ug/L	1493	Standard
	Sc	45	75343.6	2.4				ug/L	61029	Standard
	Ti	47	40945.9	1.4	**103.9053**	1.926	1.9	ug/L	35	Standard
	V	51	552663.1	1.5	**70.2550**	1.056	1.5	ug/L	1435	Standard
	Cr	52	495290.7	1.7	**63.3235**	1.419	2.2	ug/L	7635	Standard
	Cr	53	60987.5	1.0	**63.2038**	0.948	1.5	ug/L	363	Standard
	Mn	55	9453870.8	1.1	**1307.2105**	20.907	1.6	ug/L	747	Standard
	Co	59	421300.3	0.1	**61.3867**	0.363	0.6	ug/L	234	Standard
	Ni	60	119269.8	2.5	**61.2106**	1.745	2.9	ug/L	215	Standard
	Cu	65	106997.2	1.7	**55.2474**	1.210	2.2	ug/L	104	Standard
	Zn	66	101243.5	1.4	**88.7847**	1.659	1.9	ug/L	363	Standard
>	**Ge**	72	514295.3	0.5				ug/L	495142	Standard
	As	75	64887.3	0.7	**55.8337**	0.652	1.2	ug/L	-50	Standard
	Se	82	6128.3	0.4	**53.6210**	0.432	0.8	ug/L	14	Standard
	Se-1	77	4025.2	1.7	**52.4403**	1.142	2.2	ug/L	42	Standard
>	**Ga**	71	13474.3	0.3				mg/L	42	Standard
	Rb	85	148814.4	3.2				ug/L	17	Standard
	Y	89	590511.6	1.2				ug/L	417098	Standard
>	**Rh**	103	170.0	19.3				ug/L	13	Standard
	Mo	98	5249.8	1.8	**1.8847**	0.047	2.5	ug/L	11	Standard
	Ag	107	319729.4	0.9	**47.0680**	0.306	0.7	ug/L	68	Standard
	Cd	111	105379.0	1.4	**48.3587**	0.323	0.7	mg/L	2	Standard
	Cd	114	254720.6	1.7	**47.6605**	0.543	1.1	ug/L	12	Standard
>	**In**	115	612131.8	1.0				ug/L	587441	Standard
	Sn	118	2201.8	4.9	**0.2670**	0.017	6.2	ug/L	497	Standard
	Sb	123	135513.9	0.7	**27.8528**	0.396	1.4	ug/L	78	Standard
	Ba	135	276700.5	0.9	**129.7213**	0.066	0.1	ug/L	24	Standard
	Ce	140	594096.9	1.6				ug/L	25	Standard
>	**Tb**	159	982831.9	0.9				ug/L	939701	Standard
	Ho	165	12233.3	1.6				ug/L	10	Standard
	Tl	203	424338.5	0.5	**50.3728**	0.553	1.1	ug/L	4	Standard
	Tl	205	393687.9	2.5	**53.8449**	1.302	2.4	ug/L	2	Standard
	Pb	206	310823.5	0.5	**56.8661**	0.171	0.3	ug/L	240	Standard
	Pb	207	265328.4	0.5	**54.6347**	0.327	0.6	ug/L	198	Standard
	Pb	208	1090751.7	0.6	**56.8486**	0.519	0.9	ug/L	883	Standard
	U	238	409441.4	1.4	**58.2033**	0.887	1.5	ug/L	3	Standard
>	**Bi**	209	535481.7	0.7				ug/L	523984	Standard

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:20:36
Page 1

Approved: July 04, 2016

Na	23	6.7	86.6	**61.5240**	53.288	86.6	mg/L	0	Standard
Mg	24	7358.5	1.6	**7.6312**	0.188	2.5	mg/L	37	Standard
K	39	1028.4	3.5	**1.7267**	0.031	1.8	mg/L	38	Standard
Ca	43	1340.1	2.9	**66.0398**	1.115	1.7	mg/L	10	Standard
Fe	54	11628.0	2.4	**7.8474**	0.058	0.7	mg/L	179	Standard
Fe	57	3898.8	5.2	**9.6559**	0.353	3.7	mg/L	335	Standard
Sc-1	45	75343.6	2.4				mg/L	61029	Standard
Cl	35	34624.2	2.4				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	10260.1	1.8				ug/L	1547	Standard
P	31	37270.8	0.5				ug/L	15286	Standard
S	34	1845.1	3.3				ug/L	1572	Standard
Sr	88	923.4	9.7				ug/L	73	Standard
C	12	203.3	29.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	16.7	91.7				mg/L	3	Standard
Dy	164	20358.1	1.7				mg/L	6	Standard
Ho-1	165	12233.3	1.6				mg/L	10	Standard
Er	166	11334.2	2.0				mg/L	13	Standard
I	127	62766.8	7.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		134.096	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		103.868	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:20:36
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.203
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.194
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Ti 47 Upper, S, EEE	Ti	47	

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:20:36
Page 3

Approved: July 04, 2016

Mn 55 Upper, S, EEE	Mn	55
Ba 135 Upper, S, EEE	Ba	135

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:20:36
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127904SD WG574001-10
Sample Date/Time: Wednesday, June 29, 2016 17:21:31
Number of Replicates: 3
Autosampler Position: 230
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96610.6	3.3				ug/L	74699	Standard
	Be	9	1288.4	4.7	**1.1394**	0.049	4.3	ug/L	8	Standard
	Al	27	1461113.0	0.9	**7.6780**	0.187	2.4	ug/L	1493	Standard
	Sc	45	62711.5	6.5				ug/L	61029	Standard
	Ti	47	920.7	1.4	**2.1467**	0.013	0.6	ug/L	35	Standard
	V	51	13378.9	3.6	**1.4502**	0.085	5.9	ug/L	1435	Standard
	Cr	52	18691.9	1.7	**1.3185**	0.044	3.3	ug/L	7635	Standard
	Cr	53	1716.8	3.7	**1.3959**	0.091	6.5	ug/L	363	Standard
	Mn	55	207917.4	1.7	**27.5865**	0.619	2.2	ug/L	747	Standard
	Co	59	9044.0	2.3	**1.2346**	0.031	2.5	ug/L	234	Standard
	Ni	60	2824.3	2.7	**1.2771**	0.044	3.5	ug/L	215	Standard
	Cu	65	2601.9	3.8	**1.1945**	0.050	4.2	ug/L	104	Standard
	Zn	66	4831.1	2.2	**3.6537**	0.142	3.9	ug/L	363	Standard
>	**Ge**	72	533228.6	1.6				ug/L	495142	Standard
	As	75	1337.8	3.9	**1.1692**	0.061	5.2	ug/L	-50	Standard
	Se	82	140.1	2.8	**1.0475**	0.027	2.6	ug/L	14	Standard
	Se-1	77	133.7	3.0	**1.0773**	0.056	5.2	ug/L	42	Standard
>	**Ga**	71	393.3	11.4				mg/L	42	Standard
	Rb	85	3227.0	1.4				ug/L	17	Standard
	Y	89	450037.4	1.9				ug/L	417098	Standard
>	**Rh**	103	18.3	87.7				ug/L	13	Standard
	Mo	98	147.9	13.7	**0.0463**	0.009	18.8	ug/L	11	Standard
	Ag	107	7379.1	1.6	**1.0451**	0.031	3.0	ug/L	68	Standard
	Cd	111	2296.2	2.5	**1.0190**	0.058	5.7	mg/L	2	Standard
	Cd	114	5492.1	3.4	**0.9981**	0.024	2.4	ug/L	12	Standard
>	**In**	115	629018.5	3.2				ug/L	587441	Standard
	Sn	118	580.0	12.2	**-0.0020**	0.010	499.8	ug/L	497	Standard
	Sb	123	3693.7	7.1	**0.7349**	0.053	7.2	ug/L	78	Standard
	Ba	135	5980.2	1.6	**2.7136**	0.046	1.7	ug/L	24	Standard
	Ce	140	12665.3	1.3				ug/L	25	Standard
>	**Tb**	159	1007879.7	3.3				ug/L	939701	Standard
	Ho	165	265.0	7.5				ug/L	10	Standard
	Tl	203	9323.2	2.1	**1.0256**	0.019	1.8	ug/L	4	Standard
	Tl	205	8567.4	1.0	**1.0883**	0.033	3.0	ug/L	2	Standard
	Pb	206	7313.4	2.1	**1.1727**	0.042	3.6	ug/L	240	Standard
	Pb	207	6050.9	2.1	**1.0885**	0.043	4.0	ug/L	198	Standard
	Pb	208	24824.3	2.0	**1.1284**	0.015	1.4	ug/L	883	Standard
	U	238	8436.0	2.0	**1.1088**	0.013	1.2	ug/L	3	Standard
>	**Bi**	209	577014.5	3.2				ug/L	523984	Standard

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:23:48
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	200.0	10.9	**0.1869**	0.033	17.5	mg/L	37	Standard
K	39	61.7	46.1	**0.0210**	0.066	315.1	mg/L	38	Standard
Ca	43	46.7	12.4	**0.3502**	0.375	107.1	mg/L	10	Standard
Fe	54	378.5	13.7	**0.1751**	0.049	28.1	mg/L	179	Standard
Fe	57	365.0	16.4	**0.1419**	0.250	175.9	mg/L	335	Standard
> **Sc-1**	45	62711.5	6.5				mg/L	61029	Standard
Cl	35	27760.1	3.1				ug/L	20195	Standard
Kr	83	3.3	45.8				ug/L	4	Standard
Br	81	2080.1	7.5				ug/L	1547	Standard
P	31	18322.7	2.7				ug/L	15286	Standard
S	34	1913.5	2.4				ug/L	1572	Standard
Sr	88	98.3	10.6				ug/L	73	Standard
C	12	103.3	31.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	615.8	1.7				mg/L	6	Standard
Ho-1	165	265.0	7.5				mg/L	10	Standard
Er	166	296.7	10.3				mg/L	13	Standard
I	127	5109.3	32.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		129.334	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		107.692	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:23:48
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		107.078
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.121
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:23:48
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127905
Sample Date/Time: Wednesday, June 29, 2016 17:24:42
Number of Replicates: 3
Autosampler Position: 231
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98923.6	1.9				ug/L	74699	Standard
	Be	9	425.0	5.1	**0.3722**	0.020	5.4	ug/L	8	Standard
	Al	27	40729181.9	0.9	**209.2967**	3.040	1.5	ug/L	1493	Standard
	Sc	45	74965.2	4.5				ug/L	61029	Standard
	Ti	47	31002.7	2.7	**78.7018**	1.731	2.2	ug/L	35	Standard
	V	51	100570.7	0.4	**12.6354**	0.094	0.7	ug/L	1435	Standard
	Cr	52	71317.5	0.6	**8.2486**	0.021	0.3	ug/L	7635	Standard
	Cr	53	8882.6	3.6	**8.9299**	0.295	3.3	ug/L	363	Standard
	Mn	55	1183907.3	1.1	**163.7014**	1.986	1.2	ug/L	747	Standard
	Co	59	16826.7	0.9	**2.4180**	0.026	1.1	ug/L	234	Standard
	Ni	60	14706.8	1.9	**7.4444**	0.110	1.5	ug/L	215	Standard
	Cu	65	7822.7	1.1	**3.9461**	0.031	0.8	ug/L	104	Standard
	Zn	66	28811.4	1.3	**24.9594**	0.201	0.8	ug/L	363	Standard
>	**Ge**	72	513872.0	0.5				ug/L	495142	Standard
	As	75	2302.8	2.2	**2.0406**	0.054	2.6	ug/L	-50	Standard
	Se	82	647.6	3.3	**5.5477**	0.199	3.6	ug/L	14	Standard
	Se-1	77	492.7	1.9	**5.8772**	0.106	1.8	ug/L	42	Standard
>	**Ga**	71	9201.1	2.2				mg/L	42	Standard
	Rb	85	118130.4	0.8				ug/L	17	Standard
	Y	89	544220.1	0.8				ug/L	417098	Standard
>	**Rh**	103	83.3	19.3				ug/L	13	Standard
	Mo	98	4972.1	0.8	**1.6995**	0.011	0.6	ug/L	11	Standard
	Ag	107	221.0	6.7	**0.0181**	0.002	11.5	ug/L	68	Standard
	Cd	111	99.4	11.4	**0.0355**	0.005	13.9	mg/L	2	Standard
	Cd	114	258.8	15.4	**0.0440**	0.007	16.4	ug/L	12	Standard
>	**In**	115	642627.7	0.7				ug/L	587441	Standard
	Sn	118	4954.1	4.3	**0.6806**	0.035	5.1	ug/L	497	Standard
	Sb	123	697.4	10.3	**0.1324**	0.015	11.0	ug/L	78	Standard
	Ba	135	175907.0	0.9	**78.5481**	0.493	0.6	ug/L	24	Standard
	Ce	140	308696.3	0.2				ug/L	25	Standard
>	**Tb**	159	1018138.8	1.9				ug/L	939701	Standard
	Ho	165	5902.8	0.8				ug/L	10	Standard
	Tl	203	694.0	1.8	**0.0782**	0.000	0.4	ug/L	4	Standard
	Tl	205	615.0	15.7	**0.0817**	0.011	14.1	ug/L	2	Standard
	Pb	206	15714.8	1.1	**2.7233**	0.017	0.6	ug/L	240	Standard
	Pb	207	12858.1	1.8	**2.5034**	0.026	1.0	ug/L	198	Standard
	Pb	208	54664.4	0.6	**2.6954**	0.031	1.2	ug/L	883	Standard
	U	238	8998.7	0.4	**1.2376**	0.021	1.7	ug/L	3	Standard
>	**Bi**	209	551643.9	1.7				ug/L	523984	Standard

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:26:59
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**46.3980**	45.060	97.1	mg/L	0	Standard
Mg	24	7338.4	2.5	**7.6542**	0.303	4.0	mg/L	37	Standard
K	39	668.3	6.0	**1.0889**	0.064	5.9	mg/L	38	Standard
Ca	43	676.7	13.4	**32.4506**	6.121	18.9	mg/L	10	Standard
Fe	54	6512.7	2.4	**4.3601**	0.118	2.7	mg/L	179	Standard
Fe	57	2150.2	3.4	**4.8854**	0.383	7.8	mg/L	335	Standard
Sc-1	45	74965.2	4.5				mg/L	61029	Standard
Cl	35	28544.2	1.7				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	11494.4	5.7				ug/L	1547	Standard
P	31	34679.6	1.6				ug/L	15286	Standard
S	34	2073.5	2.2				ug/L	1572	Standard
Sr	88	356.7	8.9				ug/L	73	Standard
C	12	163.3	55.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	3	Standard
Dy	164	9835.4	4.4				mg/L	6	Standard
Ho-1	165	5902.8	0.8				mg/L	10	Standard
Er	166	5037.5	9.2				mg/L	13	Standard
I	127	62917.2	4.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		132.430	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		103.783	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:26:59
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.394
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.279
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:26:59
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127905
Sample Date/Time: Wednesday, June 29, 2016 17:27:54
Number of Replicates: 3
Autosampler Position: 232
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95902.7	2.3				ug/L	74699	Standard
	Be	9	26.7	54.1	**0.0312**	0.013	41.0	ug/L	8	Standard
	Al	27	866639.8	0.6	**4.5819**	0.130	2.8	ug/L	1493	Standard
	Sc	45	64717.9	1.3				ug/L	61029	Standard
	Ti	47	665.3	5.2	**1.4927**	0.099	6.6	ug/L	35	Standard
	V	51	3898.6	6.9	**0.2752**	0.030	11.1	ug/L	1435	Standard
	Cr	52	9682.4	1.6	**0.1679**	0.023	13.9	ug/L	7635	Standard
	Cr	53	670.0	10.3	**0.3316**	0.072	21.7	ug/L	363	Standard
	Mn	55	29584.3	8.9	**3.7268**	0.331	8.9	ug/L	747	Standard
	Co	59	652.3	12.4	**0.0528**	0.011	20.6	ug/L	234	Standard
	Ni	60	547.0	3.7	**0.1426**	0.010	7.1	ug/L	215	Standard
	Cu	65	341.3	5.6	**0.0636**	0.011	17.0	ug/L	104	Standard
	Zn	66	2764.3	3.0	**1.8532**	0.047	2.5	ug/L	363	Standard
>	**Ge**	72	542890.6	1.0				ug/L	495142	Standard
	As	75	11.1	320.4	**0.0688**	0.029	42.5	ug/L	-50	Standard
	Se	82	31.5	11.6	**0.1237**	0.032	26.0	ug/L	14	Standard
	Se-1	77	57.7	24.7	**0.0971**	0.173	178.1	ug/L	42	Standard
>	**Ga**	71	253.3	16.8				mg/L	42	Standard
	Rb	85	2388.5	2.4				ug/L	17	Standard
	Y	89	480185.0	1.7				ug/L	417098	Standard
>	**Rh**	103	16.7	34.6				ug/L	13	Standard
	Mo	98	144.5	13.4	**0.0432**	0.007	17.2	ug/L	11	Standard
	Ag	107	273.0	44.6	**0.0250**	0.018	70.2	ug/L	68	Standard
	Cd	111	17.1	70.4	**-0.0005**	0.005	1058.7	mg/L	2	Standard
	Cd	114	57.8	101.1	**0.0081**	0.010	129.1	ug/L	12	Standard
>	**In**	115	652419.9	1.9				ug/L	587441	Standard
	Sn	118	705.0	11.1	**0.0141**	0.014	98.8	ug/L	497	Standard
	Sb	123	247.5	32.3	**0.0437**	0.016	37.3	ug/L	78	Standard
	Ba	135	3842.5	3.8	**1.6750**	0.079	4.7	ug/L	24	Standard
	Ce	140	6778.2	3.5				ug/L	25	Standard
>	**Tb**	159	1017174.6	0.9				ug/L	939701	Standard
	Ho	165	125.0	20.8				ug/L	10	Standard
	Tl	203	149.7	88.3	**0.0148**	0.015	98.5	ug/L	4	Standard
	Tl	205	133.3	80.1	**0.0172**	0.014	79.0	ug/L	2	Standard
	Pb	206	748.0	16.8	**0.0568**	0.021	37.3	ug/L	240	Standard
	Pb	207	613.7	10.2	**0.0482**	0.012	24.5	ug/L	198	Standard
	Pb	208	2588.8	11.0	**0.0521**	0.014	26.1	ug/L	883	Standard
	U	238	255.3	28.9	**0.0295**	0.010	33.0	ug/L	3	Standard
>	**Bi**	209	573070.1	0.3				ug/L	523984	Standard

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:30:11
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.8927**	30.982	173.2	mg/L	0	Standard
Mg	24	208.3	5.5	**0.1886**	0.017	8.9	mg/L	37	Standard
K	39	35.0	14.3	**-0.0411**	0.011	27.7	mg/L	38	Standard
Ca	43	31.7	18.2	**-0.6344**	0.363	57.2	mg/L	10	Standard
Fe	54	281.0	17.0	**0.0865**	0.040	45.9	mg/L	179	Standard
Fe	57	258.3	4.5	**-0.2461**	0.039	15.8	mg/L	335	Standard
> **Sc-1**	45	64717.9	1.3				mg/L	61029	Standard
Cl	35	29516.1	1.1				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2206.8	1.5				ug/L	1547	Standard
P	31	18893.4	3.4				ug/L	15286	Standard
S	34	2140.2	7.5				ug/L	1572	Standard
Sr	88	85.0	23.5				ug/L	73	Standard
C	12	63.3	39.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	226.9	17.7				mg/L	6	Standard
Ho-1	165	125.0	20.8				mg/L	10	Standard
Er	166	133.3	38.5				mg/L	13	Standard
I	127	4113.9	27.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		128.386	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.643	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:30:11
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.061
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.368
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606128101
Sample Date/Time: Wednesday, June 29, 2016 17:31:05
Number of Replicates: 3
Autosampler Position: 233
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	95248.6	2.8				ug/L	74699	Standard
	Be	9	20.0	25.0	**0.0254**	0.004	15.8	ug/L	8	Standard
	Al	27	13040665.4	1.7	**69.5953**	1.139	1.6	ug/L	1493	Standard
	Sc	45	60164.2	2.0				ug/L	61029	Standard
	Ti	47	192.7	16.8	**0.4120**	0.087	21.1	ug/L	35	Standard
	V	51	8357.2	4.3	**0.9387**	0.048	5.1	ug/L	1435	Standard
	Cr	52	21860.8	2.2	**2.0003**	0.082	4.1	ug/L	7635	Standard
	Cr	53	28106.1	2.4	**30.8448**	0.828	2.7	ug/L	363	Standard
	Mn	55	35106.1	13.3	**5.0193**	0.674	13.4	ug/L	747	Standard
	Co	59	6553.8	6.0	**0.9799**	0.056	5.7	ug/L	234	Standard
	Ni	60	21271.9	1.7	**11.5228**	0.221	1.9	ug/L	215	Standard
	Cu	65	394730.4	1.5	**217.3087**	4.048	1.9	ug/L	104	Standard
	Zn	66	2800.3	3.3	**2.1735**	0.095	4.3	ug/L	363	Standard
>	Ge	72	483031.9	0.6				ug/L	495142	Standard
	As	75	2746.8	5.9	**2.5735**	0.152	5.9	ug/L	-50	Standard
	Se	82	892.7	0.7	**8.1998**	0.096	1.2	ug/L	14	Standard
	Se-1	77	2323.2	2.5	**31.9871**	0.998	3.1	ug/L	42	Standard
>	Ga	71	476.7	6.4				mg/L	42	Standard
	Rb	85	346107.5	1.7				ug/L	17	Standard
	Y	89	415017.7	1.3				ug/L	417098	Standard
>	Rh	103	195.0	26.0				ug/L	13	Standard
	Mo	98	2991.9	3.1	**1.1870**	0.038	3.2	ug/L	11	Standard
	Ag	107	283.3	10.3	**0.0333**	0.004	13.4	ug/L	68	Standard
	Cd	111	47.5	106.4	**0.0162**	0.025	157.7	mg/L	2	Standard
	Cd	114	106.7	47.6	**0.0200**	0.010	52.0	ug/L	12	Standard
>	In	115	552878.2	0.7				ug/L	587441	Standard
	Sn	118	3292.0	3.2	**0.5041**	0.021	4.2	ug/L	497	Standard
	Sb	123	2817.9	0.9	**0.6371**	0.010	1.5	ug/L	78	Standard
	Ba	135	115236.2	0.6	**59.8088**	0.708	1.2	ug/L	24	Standard
	Ce	140	265.0	80.7				ug/L	25	Standard
>	Tb	159	913633.5	1.5				ug/L	939701	Standard
	Ho	165	16.7	45.8				ug/L	10	Standard
	Tl	203	132.0	58.6	**0.0163**	0.011	65.5	ug/L	4	Standard
	Tl	205	136.7	39.5	**0.0217**	0.009	39.6	ug/L	2	Standard
	Pb	206	460.3	26.2	**0.0258**	0.026	100.0	ug/L	240	Standard
	Pb	207	366.3	38.5	**0.0169**	0.034	199.7	ug/L	198	Standard
	Pb	208	1427.4	18.1	**0.0116**	0.016	136.8	ug/L	883	Standard
	U	238	88.7	86.2	**0.0101**	0.013	124.4	ug/L	3	Standard
>	Bi	209	465790.3	0.8				ug/L	523984	Standard

Sample ID: L1606128101
Report Date/Time: Wednesday, June 29, 2016 17:33:22
Page 1

Approved: July 04, 2016

Na	23	8.3	69.3	**97.9562**	69.038	70.5	mg/L	0	Standard
Mg	24	359246.1	3.7	**470.2454**	9.358	2.0	mg/L	37	Standard
K	39	24917.1	2.4	**55.7677**	1.861	3.3	mg/L	38	Standard
Ca	43	1330.1	6.9	**82.7516**	6.716	8.1	mg/L	10	Standard
Fe	54	200.7	20.7	**0.0339**	0.032	94.4	mg/L	179	Standard
Fe	57	1015.0	10.0	**2.4309**	0.415	17.1	mg/L	335	Standard
> **Sc-1**	45	60164.2	2.0				mg/L	61029	Standard
Cl	35	38169.4	0.5				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	223802.3	3.7				ug/L	1547	Standard
P	31	18462.9	1.8				ug/L	15286	Standard
S	34	2156.8	3.7				ug/L	1572	Standard
Sr	88	483.3	16.2				ug/L	73	Standard
C	12	163.3	15.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	24.7				mg/L	3	Standard
Dy	164	32.2	122.0				mg/L	6	Standard
Ho-1	165	16.7	45.8				mg/L	10	Standard
Er	166	23.3	99.0				mg/L	13	Standard
I	127	949274.2	9.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		127.510	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.554	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606128101
Report Date/Time: Wednesday, June 29, 2016 17:33:22
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.116
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		88.894
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Cu 65 Upper, S, EEE	Cu	65	

Sample ID: L1606128101
Report Date/Time: Wednesday, June 29, 2016 17:33:22
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 17:34:19
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89879.3	3.0				ug/L	74699	Standard
	Be	9	54419.2	2.1	**51.4171**	2.152	4.2	ug/L	8	Standard
	Al	27	8256861.3	1.5	**46.7135**	1.756	3.8	ug/L	1493	Standard
	Sc	45	65086.2	2.0				ug/L	61029	Standard
	Ti	47	41554.9	1.2	**98.5544**	1.167	1.2	ug/L	35	Standard
	V	51	424002.3	1.3	**50.3253**	0.743	1.5	ug/L	1435	Standard
	Cr	52	412948.6	1.6	**49.1234**	1.130	2.3	ug/L	7635	Standard
	Cr	53	52872.1	1.4	**51.1473**	0.096	0.2	ug/L	363	Standard
	Mn	55	374015.3	1.0	**48.1957**	0.700	1.5	ug/L	747	Standard
	Co	59	377049.8	1.0	**51.3472**	0.802	1.6	ug/L	234	Standard
	Ni	60	101943.5	0.6	**48.8763**	0.675	1.4	ug/L	215	Standard
	Cu	65	101706.9	0.4	**49.0730**	0.490	1.0	ug/L	104	Standard
	Zn	66	60100.6	1.2	**49.0556**	0.165	0.3	ug/L	363	Standard
>	**Ge**	72	550256.1	1.3				ug/L	495142	Standard
	As	75	60702.3	2.0	**48.8234**	0.606	1.2	ug/L	-50	Standard
	Se	82	6059.9	1.5	**49.5453**	0.269	0.5	ug/L	14	Standard
	Se-1	77	4264.3	0.7	**51.9185**	0.539	1.0	ug/L	42	Standard
>	**Ga**	71	80.0	16.5				mg/L	42	Standard
	Rb	85	943.4	9.9				ug/L	17	Standard
	Y	89	456697.4	1.9				ug/L	417098	Standard
>	**Rh**	103	31.7	32.9				ug/L	13	Standard
	Mo	98	286840.3	1.0	**98.7855**	2.014	2.0	ug/L	11	Standard
	Ag	107	350242.5	1.1	**49.3132**	0.224	0.5	ug/L	68	Standard
	Cd	111	110565.3	0.7	**48.5306**	0.244	0.5	mg/L	2	Standard
	Cd	114	270182.5	3.7	**48.3546**	1.830	3.8	ug/L	12	Standard
>	**In**	115	640022.5	1.2				ug/L	587441	Standard
	Sn	118	309740.9	1.1	**48.5773**	0.779	1.6	ug/L	497	Standard
	Sb	123	247251.7	1.0	**48.6061**	0.525	1.1	ug/L	78	Standard
	Ba	135	103942.0	1.5	**46.5980**	0.695	1.5	ug/L	24	Standard
	Ce	140	81.7	18.7				ug/L	25	Standard
>	**Tb**	159	1011086.0	0.8				ug/L	939701	Standard
	Ho	165	13.3	94.4				ug/L	10	Standard
	Tl	203	438780.7	1.5	**49.1134**	0.474	1.0	ug/L	4	Standard
	Tl	205	398617.2	2.3	**51.4045**	0.522	1.0	ug/L	2	Standard
	Pb	206	284136.3	1.1	**49.0127**	0.824	1.7	ug/L	240	Standard
	Pb	207	252190.2	0.9	**48.9661**	0.976	2.0	ug/L	198	Standard
	Pb	208	1023481.0	1.1	**50.2926**	0.682	1.4	ug/L	883	Standard
	U	238	380510.7	2.1	**50.9985**	0.305	0.6	ug/L	3	Standard
>	**Bi**	209	567945.0	2.2				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:36:36
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.3524**	31.779	173.2	mg/L	0	Standard
Mg	24	4450.7	2.4	**5.3249**	0.232	4.3	mg/L	37	Standard
K	39	2040.1	4.0	**4.1161**	0.223	5.4	mg/L	38	Standard
Ca	43	101.7	19.9	**3.5019**	1.219	34.8	mg/L	10	Standard
Fe	54	6522.9	1.9	**5.0481**	0.108	2.1	mg/L	179	Standard
Fe	57	1916.8	5.3	**5.0325**	0.229	4.6	mg/L	335	Standard
Sc-1	45	65086.2	2.0				mg/L	61029	Standard
Cl	35	28870.1	1.5				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	3490.4	22.1				ug/L	1547	Standard
P	31	20260.2	3.0				ug/L	15286	Standard
S	34	2420.2	2.6				ug/L	1572	Standard
Sr	88	90.0	19.2				ug/L	73	Standard
C	12	86.7	6.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	9.5	5.0				mg/L	6	Standard
Ho-1	165	13.3	94.4				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	23970.8	64.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	102.834		
Al	27	93.427		
Sc	45			
Ti	47	98.554		
V	51	100.651		
Cr	52	98.247		
Cr	53			
Mn	55	96.391		
Co	59	102.694		
Ni	60	97.753		
Cu	65	98.146		
Zn	66	98.111		
Ge	72		111.131	
As	75	97.647		
Se	82	99.091		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:36:36
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	98.786		
Ag	107	98.626		
Cd	111	97.061		
Cd	114			
> In	115			108.951
Sn	118	97.155		
Sb	123	97.212		
Ba	135	93.196		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	98.227		
Tl	205			
Pb	206			
Pb	207			
Pb	208	100.585		
U	238	101.997		
> Bi	209			108.390
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:36:36
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 17:37:30
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90192.8	3.2				ug/L	74699	Standard
	Be	9	55.0	102.4	**0.0584**	0.051	86.9	ug/L	8	Standard
	Al	27	41697.2	123.7	**0.2160**	0.280	129.5	ug/L	1493	Standard
	Sc	45	66865.6	0.3				ug/L	61029	Standard
	Ti	47	70.3	63.0	**0.0618**	0.107	173.0	ug/L	35	Standard
	V	51	1545.5	1.7	**-0.0068**	0.004	62.6	ug/L	1435	Standard
	Cr	52	7551.9	5.0	**-0.0837**	0.044	52.6	ug/L	7635	Standard
	Cr	53	1056.7	15.9	**0.7243**	0.162	22.4	ug/L	363	Standard
	Mn	55	2100.5	33.7	**0.1299**	0.092	71.2	ug/L	747	Standard
	Co	59	307.3	14.8	**0.0056**	0.006	108.6	ug/L	234	Standard
	Ni	60	288.7	41.6	**0.0182**	0.059	321.7	ug/L	215	Standard
	Cu	65	326.7	64.3	**0.0580**	0.103	178.3	ug/L	104	Standard
	Zn	66	465.0	14.2	**-0.0620**	0.054	86.6	ug/L	363	Standard
>	**Ge**	72	536712.4	0.6				ug/L	495142	Standard
	As	75	-14.2	293.5	**0.0478**	0.034	71.9	ug/L	-50	Standard
	Se	82	26.6	3.7	**0.0859**	0.008	9.9	ug/L	14	Standard
	Se-1	77	71.3	25.4	**0.2781**	0.225	81.0	ug/L	42	Standard
>	**Ga**	71	51.7	22.3				mg/L	42	Standard
	Rb	85	111.7	73.6				ug/L	17	Standard
	Y	89	476031.3	1.1				ug/L	417098	Standard
>	**Rh**	103	26.7	47.2				ug/L	13	Standard
	Mo	98	310.3	31.4	**0.0963**	0.033	33.9	ug/L	11	Standard
	Ag	107	196.3	41.5	**0.0135**	0.011	82.4	ug/L	68	Standard
	Cd	111	44.4	76.9	**0.0108**	0.014	134.5	mg/L	2	Standard
	Cd	114	106.2	73.2	**0.0161**	0.013	83.2	ug/L	12	Standard
>	**In**	115	671001.8	0.7				ug/L	587441	Standard
	Sn	118	950.0	11.1	**0.0476**	0.015	31.2	ug/L	497	Standard
	Sb	123	463.6	17.0	**0.0827**	0.014	17.2	ug/L	78	Standard
	Ba	135	113.3	53.3	**0.0326**	0.026	80.4	ug/L	24	Standard
	Ce	140	100.0	52.9				ug/L	25	Standard
>	**Tb**	159	1042844.6	0.4				ug/L	939701	Standard
	Ho	165	21.7	53.3				ug/L	10	Standard
	Tl	203	169.7	59.1	**0.0166**	0.011	64.7	ug/L	4	Standard
	Tl	205	136.7	25.7	**0.0173**	0.004	24.8	ug/L	2	Standard
	Pb	206	350.0	21.1	**-0.0125**	0.012	95.5	ug/L	240	Standard
	Pb	207	292.7	21.7	**-0.0148**	0.012	78.6	ug/L	198	Standard
	Pb	208	1272.0	10.4	**-0.0133**	0.006	44.8	ug/L	883	Standard
	U	238	124.3	70.3	**0.0117**	0.011	96.0	ug/L	3	Standard
>	**Bi**	209	584508.5	0.7				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:39:47
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	165.0	89.4	**0.1290**	0.173	134.2	mg/L	37	Standard
K	39	35.0	14.3	**-0.0435**	0.010	22.9	mg/L	38	Standard
Ca	43	33.3	22.9	**-0.6014**	0.442	73.5	mg/L	10	Standard
Fe	54	222.3	16.4	**0.0337**	0.028	82.1	mg/L	179	Standard
Fe	57	293.3	1.0	**-0.1642**	0.011	6.9	mg/L	335	Standard
Sc-1	45	66865.6	0.3				mg/L	61029	Standard
Cl	35	28993.0	1.1				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	2536.9	15.9				ug/L	1547	Standard
P	31	21593.7	3.8				ug/L	15286	Standard
S	34	2441.9	4.5				ug/L	1572	Standard
Sr	88	71.7	39.7				ug/L	73	Standard
C	12	43.3	53.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	2.2	286.4				mg/L	6	Standard
Ho-1	165	21.7	53.3				mg/L	10	Standard
Er	166	23.3	65.5				mg/L	13	Standard
I	127	6326.4	22.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		108.396	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:39:47
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.225
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.551
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:39:47
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129403
Sample Date/Time: Wednesday, June 29, 2016 17:40:43
Number of Replicates: 3
Autosampler Position: 234
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	91138.4	2.1				ug/L	74699	Standard
	Be	9	18.3	31.5	**0.0246**	0.005	20.1	ug/L	8	Standard
	Al	27	19782.9	2.5	**0.0970**	0.003	3.0	ug/L	1493	Standard
	Sc	45	69032.0	3.0				ug/L	61029	Standard
	Ti	47	54.0	3.2	**0.0194**	0.002	12.6	ug/L	35	Standard
	V	51	1966.3	6.5	**0.0390**	0.011	28.1	ug/L	1435	Standard
	Cr	52	10805.9	1.1	**0.2915**	0.043	14.7	ug/L	7635	Standard
	Cr	53	2143.5	1.8	**1.7623**	0.077	4.4	ug/L	363	Standard
	Mn	55	3844.5	2.2	**0.3500**	0.018	5.0	ug/L	747	Standard
	Co	59	327.3	8.7	**0.0074**	0.003	41.0	ug/L	234	Standard
	Ni	60	356.0	5.4	**0.0475**	0.008	16.5	ug/L	215	Standard
	Cu	65	696.0	2.6	**0.2337**	0.014	6.0	ug/L	104	Standard
	Zn	66	2205.8	3.0	**1.3661**	0.024	1.8	ug/L	363	Standard
>	**Ge**	72	549152.6	2.2				ug/L	495142	Standard
	As	75	-29.6	63.4	**0.0356**	0.015	42.9	ug/L	-50	Standard
	Se	82	23.9	20.2	**0.0586**	0.040	68.9	ug/L	14	Standard
	Se-1	77	94.7	7.0	**0.5478**	0.105	19.2	ug/L	42	Standard
>	**Ga**	71	51.7	14.8				mg/L	42	Standard
	Rb	85	138.3	25.6				ug/L	17	Standard
	Y	89	465979.4	3.6				ug/L	417098	Standard
>	**Rh**	103	13.3	21.7				ug/L	13	Standard
	Mo	98	5139.1	1.4	**1.7220**	0.056	3.3	ug/L	11	Standard
	Ag	107	110.0	6.9	**0.0022**	0.001	48.4	ug/L	68	Standard
	Cd	111	15.0	26.6	**-0.0015**	0.002	117.9	mg/L	2	Standard
	Cd	114	55.2	20.3	**0.0075**	0.002	25.2	ug/L	12	Standard
>	**In**	115	655841.8	1.8				ug/L	587441	Standard
	Sn	118	1048.4	10.6	**0.0659**	0.015	22.6	ug/L	497	Standard
	Sb	123	219.8	14.7	**0.0379**	0.005	14.5	ug/L	78	Standard
	Ba	135	307.7	5.8	**0.1187**	0.008	7.0	ug/L	24	Standard
	Ce	140	120.0	27.3				ug/L	25	Standard
>	**Tb**	159	1025647.2	2.3				ug/L	939701	Standard
	Ho	165	10.0	50.0				ug/L	10	Standard
	Tl	203	102.7	26.3	**0.0095**	0.003	29.3	ug/L	4	Standard
	Tl	205	93.3	17.2	**0.0120**	0.002	17.6	ug/L	2	Standard
	Pb	206	348.0	5.5	**-0.0123**	0.003	25.6	ug/L	240	Standard
	Pb	207	291.0	7.0	**-0.0146**	0.003	21.1	ug/L	198	Standard
	Pb	208	1220.0	1.5	**-0.0152**	0.002	10.4	ug/L	883	Standard
	U	238	43.7	3.5	**0.0013**	0.000	20.2	ug/L	3	Standard
>	**Bi**	209	579311.4	1.7				ug/L	523984	Standard

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:43:00
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	70.0	14.3	**0.0151**	0.014	91.8	mg/L	37	Standard
K	39	41.7	36.7	**-0.0323**	0.032	98.5	mg/L	38	Standard
Ca	43	23.3	24.7	**-1.2220**	0.307	25.1	mg/L	10	Standard
Fe	54	180.7	18.1	**-0.0033**	0.021	642.1	mg/L	179	Standard
Fe	57	228.3	18.9	**-0.3899**	0.112	28.8	mg/L	335	Standard
> **Sc-1**	45	69032.0	3.0				mg/L	61029	Standard
Cl	35	29393.1	0.1				ug/L	20195	Standard
Kr	83	3.7	41.7				ug/L	4	Standard
Br	81	2280.2	16.1				ug/L	1547	Standard
P	31	20889.4	5.8				ug/L	15286	Standard
S	34	2440.2	2.2				ug/L	1572	Standard
Sr	88	78.3	9.8				ug/L	73	Standard
C	12	90.0	19.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	5.2	126.6				mg/L	6	Standard
Ho-1	165	10.0	50.0				mg/L	10	Standard
Er	166	30.0	66.7				mg/L	13	Standard
I	127	4447.3	11.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		122.008	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.908	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:43:00
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.644
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.559
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129403
Sample Date/Time: Wednesday, June 29, 2016 17:43:55
Number of Replicates: 3
Autosampler Position: 235
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89753.4	1.3				ug/L	74699	Standard
	Be	9	13.3	43.3	**0.0203**	0.006	27.7	ug/L	8	Standard
	Al	27	7761551.3	2.6	**43.9444**	1.176	2.7	ug/L	1493	Standard
	Sc	45	64855.2	1.4				ug/L	61029	Standard
	Ti	47	206.0	13.1	**0.3915**	0.063	16.1	ug/L	35	Standard
	V	51	6651.8	1.7	**0.6159**	0.016	2.5	ug/L	1435	Standard
	Cr	52	10534.3	3.3	**0.2860**	0.045	15.8	ug/L	7635	Standard
	Cr	53	2358.5	8.3	**2.0219**	0.206	10.2	ug/L	363	Standard
	Mn	55	48108.9	1.2	**6.2182**	0.039	0.6	ug/L	747	Standard
	Co	59	779.0	1.7	**0.0714**	0.002	2.2	ug/L	234	Standard
	Ni	60	2066.1	2.9	**0.8928**	0.027	3.0	ug/L	215	Standard
	Cu	65	1897.5	0.8	**0.8356**	0.011	1.4	ug/L	104	Standard
	Zn	66	4161.9	4.0	**3.0555**	0.122	4.0	ug/L	363	Standard
>	**Ge**	72	537411.3	0.6				ug/L	495142	Standard
	As	75	1930.1	2.9	**1.6472**	0.048	2.9	ug/L	-50	Standard
	Se	82	65.9	12.2	**0.4151**	0.071	17.1	ug/L	14	Standard
	Se-1	77	143.0	7.4	**1.1820**	0.141	12.0	ug/L	42	Standard
>	**Ga**	71	281.7	11.8				mg/L	42	Standard
	Rb	85	9322.9	0.6				ug/L	17	Standard
	Y	89	483828.6	0.3				ug/L	417098	Standard
>	**Rh**	103	65.0	35.3				ug/L	13	Standard
	Mo	98	2420.2	2.2	**0.8025**	0.008	1.0	ug/L	11	Standard
	Ag	107	86.7	9.0	**-0.0011**	0.001	98.3	ug/L	68	Standard
	Cd	111	21.1	19.4	**0.0011**	0.002	172.0	mg/L	2	Standard
	Cd	114	57.2	8.0	**0.0078**	0.001	8.8	ug/L	12	Standard
>	**In**	115	659947.0	1.4				ug/L	587441	Standard
	Sn	118	1036.7	2.5	**0.0632**	0.002	2.8	ug/L	497	Standard
	Sb	123	514.2	4.3	**0.0939**	0.005	5.0	ug/L	78	Standard
	Ba	135	42292.6	1.8	**18.3786**	0.354	1.9	ug/L	24	Standard
	Ce	140	906.7	5.0				ug/L	25	Standard
>	**Tb**	159	1004743.5	1.0				ug/L	939701	Standard
	Ho	165	30.0	33.3				ug/L	10	Standard
	Tl	203	321.7	3.2	**0.0348**	0.001	3.5	ug/L	4	Standard
	Tl	205	263.3	19.0	**0.0346**	0.006	17.3	ug/L	2	Standard
	Pb	206	528.0	6.8	**0.0212**	0.005	24.5	ug/L	240	Standard
	Pb	207	478.3	8.2	**0.0244**	0.009	35.8	ug/L	198	Standard
	Pb	208	1973.7	3.1	**0.0244**	0.005	18.6	ug/L	883	Standard
	U	238	903.7	2.1	**0.1184**	0.002	1.8	ug/L	3	Standard
>	**Bi**	209	560080.8	1.7				ug/L	523984	Standard

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:46:12
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	9985.0	3.9	**12.0618**	0.357	3.0	mg/L	37	Standard
K	39	420.0	7.4	**0.7593**	0.061	8.1	mg/L	38	Standard
Ca	43	263.3	2.9	**13.1350**	0.654	5.0	mg/L	10	Standard
Fe	54	210.7	11.6	**0.0297**	0.017	57.3	mg/L	179	Standard
Fe	57	393.3	9.5	**0.1838**	0.119	64.8	mg/L	335	Standard
> **Sc-1**	45	64855.2	1.4				mg/L	61029	Standard
Cl	35	32877.0	2.8				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	8902.6	4.1				ug/L	1547	Standard
P	31	19829.6	4.1				ug/L	15286	Standard
S	34	2240.2	4.2				ug/L	1572	Standard
Sr	88	221.7	14.7				ug/L	73	Standard
C	12	126.7	4.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	64.1	36.5				mg/L	6	Standard
Ho-1	165	30.0	33.3				mg/L	10	Standard
Er	166	53.3	21.7				mg/L	13	Standard
I	127	24610.0	6.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		120.154	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.537	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:46:12
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.343
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.889
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129601
Sample Date/Time: Wednesday, June 29, 2016 17:47:06
Number of Replicates: 3
Autosampler Position: 236
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89414.7	0.8				ug/L	74699	Standard
	Be	9	18.3	41.7	**0.0250**	0.007	28.7	ug/L	8	Standard
	Al	27	159581.9	4.0	**0.8936**	0.029	3.3	ug/L	1493	Standard
	Sc	45	67088.2	1.1				ug/L	61029	Standard
	Ti	47	35.7	1.6	**-0.0268**	0.002	7.8	ug/L	35	Standard
	V	51	2172.9	6.2	**0.0559**	0.017	29.8	ug/L	1435	Standard
	Cr	52	9358.6	1.2	**0.0800**	0.026	32.1	ug/L	7635	Standard
	Cr	53	761.7	7.6	**0.3908**	0.063	16.2	ug/L	363	Standard
	Mn	55	2132.8	0.4	**0.1197**	0.005	4.4	ug/L	747	Standard
	Co	59	232.7	3.6	**-0.0065**	0.001	22.5	ug/L	234	Standard
	Ni	60	263.3	7.3	**-0.0009**	0.011	1213.2	ug/L	215	Standard
	Cu	65	215.7	8.8	**-0.0024**	0.009	395.1	ug/L	104	Standard
	Zn	66	2535.5	4.0	**1.5735**	0.060	3.8	ug/L	363	Standard
>	**Ge**	72	566628.0	1.5				ug/L	495142	Standard
	As	75	0.7	2838.5	**0.0602**	0.015	24.5	ug/L	-50	Standard
	Se	82	16.5	13.6	**-0.0062**	0.020	325.1	ug/L	14	Standard
	Se-1	77	57.7	30.4	**0.0671**	0.205	305.9	ug/L	42	Standard
>	**Ga**	71	65.0	35.3				mg/L	42	Standard
	Rb	85	218.3	8.7				ug/L	17	Standard
	Y	89	474347.0	1.6				ug/L	417098	Standard
>	**Rh**	103	16.7	45.8				ug/L	13	Standard
	Mo	98	108.8	9.6	**0.0309**	0.003	10.7	ug/L	11	Standard
	Ag	107	170.0	11.0	**0.0105**	0.003	26.1	ug/L	68	Standard
	Cd	111	6.5	32.0	**-0.0051**	0.001	17.0	mg/L	2	Standard
	Cd	114	10.6	76.1	**-0.0002**	0.001	576.3	ug/L	12	Standard
>	**In**	115	655393.1	1.0				ug/L	587441	Standard
	Sn	118	536.7	6.3	**-0.0123**	0.006	49.5	ug/L	497	Standard
	Sb	123	179.2	10.0	**0.0302**	0.004	12.2	ug/L	78	Standard
	Ba	135	901.0	1.9	**0.3786**	0.012	3.1	ug/L	24	Standard
	Ce	140	38.3	52.7				ug/L	25	Standard
>	**Tb**	159	1019773.4	1.2				ug/L	939701	Standard
	Ho	165	10.0	50.0				ug/L	10	Standard
	Tl	203	34.7	13.6	**0.0021**	0.001	27.9	ug/L	4	Standard
	Tl	205	33.3	52.7	**0.0044**	0.002	51.3	ug/L	2	Standard
	Pb	206	333.0	5.0	**-0.0144**	0.003	23.5	ug/L	240	Standard
	Pb	207	282.3	6.2	**-0.0157**	0.004	25.6	ug/L	198	Standard
	Pb	208	1187.0	2.0	**-0.0164**	0.002	10.6	ug/L	883	Standard
	U	238	27.0	25.9	**-0.0009**	0.001	108.4	ug/L	3	Standard
>	**Bi**	209	574894.3	1.4				ug/L	523984	Standard

Sample ID: L1606129601
Report Date/Time: Wednesday, June 29, 2016 17:49:23
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.3542**	30.050	173.2	mg/L	0	Standard
Mg	24	255.0	18.0	**0.2345**	0.054	23.2	mg/L	37	Standard
K	39	41.7	36.7	**-0.0306**	0.030	98.0	mg/L	38	Standard
Ca	43	31.7	18.2	**-0.7022**	0.343	48.8	mg/L	10	Standard
Fe	54	182.6	4.3	**0.0025**	0.004	177.8	mg/L	179	Standard
Fe	57	238.3	19.9	**-0.3363**	0.155	46.0	mg/L	335	Standard
Sc-1	45	67088.2	1.1				mg/L	61029	Standard
Cl	35	32679.2	0.8				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	2083.5	4.4				ug/L	1547	Standard
P	31	19699.5	2.2				ug/L	15286	Standard
S	34	2248.5	3.2				ug/L	1572	Standard
Sr	88	70.0	56.7				ug/L	73	Standard
C	12	60.0	76.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	42.4	37.5				mg/L	6	Standard
Ho-1	165	10.0	50.0				mg/L	10	Standard
Er	166	20.0	86.6				mg/L	13	Standard
I	127	3452.1	27.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		119.700	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.437	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606129601
Report Date/Time: Wednesday, June 29, 2016 17:49:23
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.567
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.716
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Wednesday, June 29, 2016 17:50:19
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88962.1	0.9				ug/L	74699	Standard
	Be	9	13.3	43.3	**0.0204**	0.006	27.5	ug/L	8	Standard
	Al	27	8687475.6	2.0	**49.6312**	1.401	2.8	ug/L	1493	Standard
	Sc	45	67439.8	1.2				ug/L	61029	Standard
	Ti	47	49495.1	1.8	**118.6714**	2.776	2.3	ug/L	35	Standard
	V	51	2497.1	11.5	**0.1050**	0.033	31.3	ug/L	1435	Standard
	Cr	52	8893.6	0.7	**0.0677**	0.012	17.4	ug/L	7635	Standard
	Cr	53	8300.6	3.8	**7.8359**	0.175	2.2	ug/L	363	Standard
	Mn	55	1506.4	1.9	**0.0487**	0.003	6.1	ug/L	747	Standard
	Co	59	411.3	2.2	**0.0194**	0.002	11.4	ug/L	234	Standard
	Ni	60	986.0	7.4	**0.3548**	0.029	8.2	ug/L	215	Standard
	Cu	65	754.7	3.5	**0.2650**	0.009	3.4	ug/L	104	Standard
	Zn	66	1997.5	3.6	**1.2085**	0.053	4.4	ug/L	363	Standard
>	**Ge**	72	544457.7	1.7				ug/L	495142	Standard
	As	75	-43.8	164.6	**0.0238**	0.058	245.3	ug/L	-50	Standard
	Se	82	18.8	19.8	**0.0181**	0.034	186.0	ug/L	14	Standard
	Se-1	77	455.0	4.4	**5.0437**	0.245	4.9	ug/L	42	Standard
>	**Ga**	71	186.7	17.4				mg/L	42	Standard
	Rb	85	1130.0	9.7				ug/L	17	Standard
	Y	89	464215.3	1.7				ug/L	417098	Standard
>	**Rh**	103	10.0	100.0				ug/L	13	Standard
	Mo	98	300673.2	0.9	**102.5347**	1.457	1.4	ug/L	11	Standard
	Ag	107	252.3	6.3	**0.0223**	0.003	12.6	ug/L	68	Standard
	Cd	111	-194.8	14.0	**-0.0926**	0.012	12.9	mg/L	2	Standard
	Cd	114	577.1	3.2	**0.1002**	0.004	4.2	ug/L	12	Standard
>	**In**	115	646355.5	1.8				ug/L	587441	Standard
	Sn	118	791.7	5.4	**0.0285**	0.006	21.5	ug/L	497	Standard
	Sb	123	195.6	12.5	**0.0338**	0.004	11.9	ug/L	78	Standard
	Ba	135	712.0	4.0	**0.3001**	0.008	2.7	ug/L	24	Standard
	Ce	140	246.7	7.1				ug/L	25	Standard
>	**Tb**	159	1024202.6	1.1				ug/L	939701	Standard
	Ho	165	23.3	61.9				ug/L	10	Standard
	Tl	203	122.3	9.8	**0.0119**	0.001	11.8	ug/L	4	Standard
	Tl	205	90.0	11.1	**0.0118**	0.001	10.5	ug/L	2	Standard
	Pb	206	533.7	2.0	**0.0208**	0.002	10.5	ug/L	240	Standard
	Pb	207	414.7	3.2	**0.0105**	0.002	21.4	ug/L	198	Standard
	Pb	208	1800.4	3.2	**0.0144**	0.003	21.4	ug/L	883	Standard
	U	238	18.7	20.3	**-0.0019**	0.001	26.8	ug/L	3	Standard
>	**Bi**	209	568876.5	0.5				ug/L	523984	Standard

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 17:52:36
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	11507.7	5.3	**13.3841**	0.868	6.5	mg/L	37	Standard
K	39	2163.5	5.0	**4.2129**	0.203	4.8	mg/L	38	Standard
Ca	43	273.3	20.1	**13.1286**	3.349	25.5	mg/L	10	Standard
Fe	54	17120.6	2.5	**12.9991**	0.385	3.0	mg/L	179	Standard
Fe	57	4470.7	3.8	**12.6745**	0.387	3.1	mg/L	335	Standard
> **Sc-1**	45	67439.8	1.2				mg/L	61029	Standard
Cl	35	28777.3	2.0				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	2086.8	10.9				ug/L	1547	Standard
P	31	20233.5	1.2				ug/L	15286	Standard
S	34	2398.5	7.6				ug/L	1572	Standard
Sr	88	63.3	35.6				ug/L	73	Standard
C	12	173.3	8.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	80.0	69.6				mg/L	3	Standard
Dy	164	11.6	52.9				mg/L	6	Standard
Ho-1	165	23.3	61.9				mg/L	10	Standard
Er	166	36.7	56.8				mg/L	13	Standard
I	127	2830.3	21.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27	0.993		
Sc	45			
Ti	47	118.671		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.960	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 17:52:36
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.535		
Ag	107			
Cd	111			
Cd	114			
In	115			110.029
Sn	118			
Sb	123			
Ba	135			
Ce	140			
Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
Bi	209			108.567
Na	23	0.040		
Mg	24	267.683		
K	39	84.259		
Ca	43	87.524		
Fe	54	103.992		
Fe	57	101.396		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Na	23	
QC Std 4	Mg	24	

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 17:52:36
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Wednesday, June 29, 2016 17:53:31
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88668.7	1.6				ug/L	74699	Standard
	Be	9	115505.6	3.3	**110.5278**	2.817	2.5	ug/L	8	Standard
	Al	27	8938177.8	2.5	**51.2202**	0.471	0.9	ug/L	1493	Standard
	Sc	45	65087.9	3.1				ug/L	61029	Standard
	Ti	47	50197.4	1.2	**122.5394**	1.491	1.2	ug/L	35	Standard
	V	51	859522.7	1.3	**105.1949**	1.380	1.3	ug/L	1435	Standard
	Cr	52	831098.2	2.0	**102.8267**	1.943	1.9	ug/L	7635	Standard
	Cr	53	106911.0	2.3	**106.7884**	2.096	2.0	ug/L	363	Standard
	Mn	55	780341.4	1.6	**103.6433**	1.453	1.4	ug/L	747	Standard
	Co	59	785536.0	0.5	**110.1234**	0.547	0.5	ug/L	234	Standard
	Ni	60	203753.2	2.0	**100.6522**	1.808	1.8	ug/L	215	Standard
	Cu	65	201732.1	0.8	**100.2703**	0.833	0.8	ug/L	104	Standard
	Zn	66	122540.9	0.7	**103.4323**	0.376	0.4	ug/L	363	Standard
>	**Ge**	72	534677.6	0.4				ug/L	495142	Standard
	As	75	124456.8	0.2	**102.9556**	0.389	0.4	ug/L	-50	Standard
	Se	82	12596.7	0.4	**106.1488**	0.433	0.4	ug/L	14	Standard
	Se-1	77	8794.6	0.8	**110.8893**	1.382	1.2	ug/L	42	Standard
>	**Ga**	71	163.3	10.8				mg/L	42	Standard
	Rb	85	966.7	4.8				ug/L	17	Standard
	Y	89	476525.6	1.7				ug/L	417098	Standard
>	**Rh**	103	56.7	53.9				ug/L	13	Standard
	Mo	98	309227.9	0.6	**104.8818**	1.132	1.1	ug/L	11	Standard
	Ag	107	592523.9	2.1	**82.1821**	1.932	2.4	ug/L	68	Standard
	Cd	111	228625.7	1.0	**98.8467**	0.904	0.9	mg/L	2	Standard
	Cd	114	551600.1	1.5	**97.2335**	1.523	1.6	ug/L	12	Standard
>	**In**	115	649797.8	0.6				ug/L	587441	Standard
	Sn	118	1576.7	4.2	**0.1493**	0.011	7.5	ug/L	497	Standard
	Sb	123	495027.3	1.9	**95.8585**	2.358	2.5	ug/L	78	Standard
	Ba	135	210763.3	1.3	**93.0830**	1.687	1.8	ug/L	24	Standard
	Ce	140	70.0	24.7				ug/L	25	Standard
>	**Tb**	159	1005734.4	0.5				ug/L	939701	Standard
	Ho	165	11.7	49.5				ug/L	10	Standard
	Tl	203	893894.1	1.2	**101.2389**	0.317	0.3	ug/L	4	Standard
	Tl	205	815034.2	0.6	**106.3608**	0.654	0.6	ug/L	2	Standard
	Pb	206	583528.7	0.9	**101.9143**	0.243	0.2	ug/L	240	Standard
	Pb	207	499626.1	1.2	**98.2110**	0.372	0.4	ug/L	198	Standard
	Pb	208	2105185.2	0.8	**104.7451**	0.431	0.4	ug/L	883	Standard
	U	238	779429.7	1.7	**105.7078**	0.634	0.6	ug/L	3	Standard
>	**Bi**	209	561247.5	1.2				ug/L	523984	Standard

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.1880**	31.494	173.2	mg/L	0	Standard
Mg	24	12009.8	3.1	**14.4710**	0.020	0.1	mg/L	37	Standard
K	39	2150.2	4.2	**4.3456**	0.256	5.9	mg/L	38	Standard
Ca	43	251.7	12.1	**12.4083**	2.087	16.8	mg/L	10	Standard
Fe	54	17836.2	2.6	**14.0504**	0.592	4.2	mg/L	179	Standard
Fe	57	4537.3	4.4	**13.3981**	0.845	6.3	mg/L	335	Standard
Sc-1	45	65087.9	3.1				mg/L	61029	Standard
Cl	35	27559.7	2.9				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	1986.8	4.8				ug/L	1547	Standard
P	31	19052.0	2.9				ug/L	15286	Standard
S	34	2286.8	6.3				ug/L	1572	Standard
Sr	88	50.0	10.0				ug/L	73	Standard
C	12	120.0	54.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	6.0	200.7				mg/L	6	Standard
Ho-1	165	11.7	49.5				mg/L	10	Standard
Er	166	13.3	86.6				mg/L	13	Standard
I	127	2065.1	25.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	110.528		
Al	27	1.024		
Sc	45			
Ti	47	122.539		
V	51	105.195		
Cr	52	102.827		
Cr	53			
Mn	55	103.643		
Co	59	110.123		
Ni	60	100.652		
Cu	65	100.270		
Zn	66	103.432		
Ge	72		107.985	
As	75	102.956		
Se	82	106.149		
Se-1	77			
Ga	71			

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	104.882		
Ag	107	82.182		
Cd	111	98.847		
Cd	114			
In	115		110.615	
Sn	118			
Sb	123	95.858		
Ba	135	93.083		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.239		
Tl	205			
Pb	206			
Pb	207			
Pb	208	104.745		
U	238	105.708		
Bi	209		107.111	
Na	23	145.504		
Mg	24	289.419		
K	39	86.911		
Ca	43	82.722		
Fe	54	112.403		
Fe	57	107.185		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Ti	47	
QC Std 5	Na	23	

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 3

Approved: July 04, 2016

QC Std 5	Mg	24

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 17:56:44
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	88001.7	3.8				ug/L	74699	Standard
	Be	9	58491.1	3.4	**56.4121**	0.807	1.4	ug/L	8	Standard
	Al	27	8741223.1	1.9	**50.5126**	1.685	3.3	ug/L	1493	Standard
	Sc	45	66566.0	2.9				ug/L	61029	Standard
	Ti	47	44065.3	0.6	**105.8464**	2.393	2.3	ug/L	35	Standard
	V	51	439728.9	0.9	**52.8674**	1.388	2.6	ug/L	1435	Standard
	Cr	52	434693.1	0.1	**52.4281**	0.919	1.8	ug/L	7635	Standard
	Cr	53	53735.1	1.4	**52.6541**	1.206	2.3	ug/L	363	Standard
	Mn	55	394255.2	1.0	**51.4522**	0.719	1.4	ug/L	747	Standard
	Co	59	403494.7	0.9	**55.6456**	1.136	2.0	ug/L	234	Standard
	Ni	60	106493.3	0.2	**51.7076**	0.770	1.5	ug/L	215	Standard
	Cu	65	105229.8	0.3	**51.4208**	1.011	2.0	ug/L	104	Standard
	Zn	66	61955.4	1.5	**51.2322**	1.156	2.3	ug/L	363	Standard
>	Ge	72	543443.2	1.7				ug/L	495142	Standard
	As	75	62757.8	1.3	**51.1208**	1.290	2.5	ug/L	-50	Standard
	Se	82	6411.0	0.6	**53.0945**	1.065	2.0	ug/L	14	Standard
	Se-1	77	4401.6	1.7	**54.2995**	1.443	2.7	ug/L	42	Standard
>	Ga	71	61.7	12.4				mg/L	42	Standard
	Rb	85	943.4	4.4				ug/L	17	Standard
	Y	89	472814.1	2.0				ug/L	417098	Standard
>	Rh	103	31.7	36.5				ug/L	13	Standard
	Mo	98	301859.8	0.5	**101.5547**	1.508	1.5	ug/L	11	Standard
	Ag	107	366041.5	1.2	**50.3642**	1.561	3.1	ug/L	68	Standard
	Cd	111	117544.4	1.3	**50.4048**	0.748	1.5	mg/L	2	Standard
	Cd	114	286950.2	2.0	**50.1688**	0.955	1.9	ug/L	12	Standard
>	In	115	655193.6	1.9				ug/L	587441	Standard
	Sn	118	333959.0	1.2	**51.1818**	1.510	3.0	ug/L	497	Standard
	Sb	123	261510.0	1.1	**50.2357**	1.515	3.0	ug/L	78	Standard
	Ba	135	110572.3	0.3	**48.4318**	0.852	1.8	ug/L	24	Standard
	Ce	140	81.7	46.0				ug/L	25	Standard
>	Tb	159	1009652.9	1.9				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	466012.5	0.6	**52.5145**	1.216	2.3	ug/L	4	Standard
	Tl	205	428260.1	0.9	**55.6109**	1.642	3.0	ug/L	2	Standard
	Pb	206	294247.5	0.8	**51.0977**	1.194	2.3	ug/L	240	Standard
	Pb	207	260542.8	0.7	**50.9223**	1.007	2.0	ug/L	198	Standard
	Pb	208	1081629.9	0.8	**53.5152**	1.532	2.9	ug/L	883	Standard
	U	238	403586.7	0.9	**54.4635**	1.402	2.6	ug/L	3	Standard
>	Bi	209	564239.9	2.1				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:59:01
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4549.0	2.0	**5.3196**	0.066	1.2	mg/L	37	Standard
K	39	2091.8	2.6	**4.1250**	0.021	0.5	mg/L	38	Standard
Ca	43	106.7	9.8	**3.6589**	0.645	17.6	mg/L	10	Standard
Fe	54	6853.2	4.1	**5.1952**	0.357	6.9	mg/L	179	Standard
Fe	57	1901.8	4.2	**4.8507**	0.081	1.7	mg/L	335	Standard
> **Sc-1**	45	66566.0	2.9				mg/L	61029	Standard
Cl	35	28481.4	2.6				ug/L	20195	Standard
Kr	83	3.3	75.5				ug/L	4	Standard
Br	81	1840.1	18.3				ug/L	1547	Standard
P	31	20427.1	5.4				ug/L	15286	Standard
S	34	2438.5	6.3				ug/L	1572	Standard
Sr	88	80.0	32.5				ug/L	73	Standard
C	12	76.7	7.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	6.5	86.7				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	2150.2	26.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	112.824		
Al	27	101.025		
Sc	45			
Ti	47	105.846		
V	51	105.735		
Cr	52	104.856		
Cr	53			
Mn	55	102.904		
Co	59	111.291		
Ni	60	103.415		
Cu	65	102.842		
Zn	66	102.464		
> Ge	72		109.755	
As	75	102.242		
Se	82	106.189		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:59:01
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.555		
Ag	107	100.728		
Cd	111	100.810		
Cd	114			
In	115		111.533	
Sn	118	102.364		
Sb	123	100.471		
Ba	135	96.864		
Ce	140			
Tb	159			
Ho	165			
Tl	203	105.029		
Tl	205			
Pb	206			
Pb	207			
Pb	208	107.030		
U	238	108.927		
Bi	209		107.683	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:59:01
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 17:59:55
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86353.7	1.4				ug/L	74699	Standard
	Be	9	83.3	67.4	**0.0891**	0.054	60.5	ug/L	8	Standard
	Al	27	42873.5	92.7	**0.2376**	0.231	97.1	ug/L	1493	Standard
	Sc	45	65305.4	1.2				ug/L	61029	Standard
	Ti	47	79.7	51.6	**0.0788**	0.097	123.4	ug/L	35	Standard
	V	51	1461.2	4.4	**-0.0225**	0.008	34.9	ug/L	1435	Standard
	Cr	52	7626.9	6.9	**-0.1029**	0.066	64.1	ug/L	7635	Standard
	Cr	53	760.0	14.8	**0.4057**	0.109	26.8	ug/L	363	Standard
	Mn	55	4546.9	137.2	**0.4388**	0.807	183.9	ug/L	747	Standard
	Co	59	626.0	93.3	**0.0478**	0.080	167.0	ug/L	234	Standard
	Ni	60	409.0	51.9	**0.0715**	0.102	142.7	ug/L	215	Standard
	Cu	65	272.7	44.7	**0.0274**	0.059	215.3	ug/L	104	Standard
	Zn	66	514.0	23.7	**-0.0335**	0.100	300.2	ug/L	363	Standard
>	**Ge**	72	553485.6	1.3				ug/L	495142	Standard
	As	75	-6.7	474.8	**0.0545**	0.025	45.9	ug/L	-50	Standard
	Se	82	20.4	10.3	**0.0278**	0.015	55.0	ug/L	14	Standard
	Se-1	77	66.3	4.4	**0.1906**	0.046	24.0	ug/L	42	Standard
>	**Ga**	71	50.0	10.0				mg/L	42	Standard
	Rb	85	45.0	48.4				ug/L	17	Standard
	Y	89	474090.9	1.7				ug/L	417098	Standard
>	**Rh**	103	18.3	41.7				ug/L	13	Standard
	Mo	98	533.3	54.8	**0.1739**	0.097	56.0	ug/L	11	Standard
	Ag	107	205.7	25.6	**0.0154**	0.007	45.9	ug/L	68	Standard
	Cd	111	43.3	79.3	**0.0106**	0.015	137.7	mg/L	2	Standard
	Cd	114	160.8	66.0	**0.0260**	0.018	70.8	ug/L	12	Standard
>	**In**	115	653140.9	0.6				ug/L	587441	Standard
	Sn	118	840.0	19.7	**0.0347**	0.026	74.7	ug/L	497	Standard
	Sb	123	1054.2	24.1	**0.1991**	0.050	25.2	ug/L	78	Standard
	Ba	135	82.0	54.8	**0.0199**	0.020	97.9	ug/L	24	Standard
	Ce	140	78.3	62.0				ug/L	25	Standard
>	**Tb**	159	1021325.7	0.5				ug/L	939701	Standard
	Ho	165	20.0	66.1				ug/L	10	Standard
	Tl	203	163.0	50.8	**0.0163**	0.009	56.1	ug/L	4	Standard
	Tl	205	543.4	156.5	**0.0696**	0.109	156.0	ug/L	2	Standard
	Pb	206	336.7	16.7	**-0.0136**	0.010	70.5	ug/L	240	Standard
	Pb	207	309.7	12.8	**-0.0104**	0.008	73.2	ug/L	198	Standard
	Pb	208	2162.1	83.2	**0.0313**	0.088	280.1	ug/L	883	Standard
	U	238	524.4	153.4	**0.0652**	0.107	163.8	ug/L	3	Standard
>	**Bi**	209	573163.8	0.1				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:02:12
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.8448**	30.899	173.2	mg/L	0	Standard
Mg	24	123.3	127.6	**0.0829**	0.188	227.1	mg/L	37	Standard
K	39	30.0	66.7	**-0.0524**	0.041	77.8	mg/L	38	Standard
Ca	43	21.7	35.3	**-1.2409**	0.467	37.7	mg/L	10	Standard
Fe	54	185.6	27.6	**0.0086**	0.039	460.9	mg/L	179	Standard
Fe	57	268.3	10.3	**-0.2224**	0.079	35.7	mg/L	335	Standard
> **Sc-1**	45	65305.4	1.2				mg/L	61029	Standard
Cl	35	28806.0	1.7				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1900.1	3.3				ug/L	1547	Standard
P	31	20240.2	2.7				ug/L	15286	Standard
S	34	2435.2	7.1				ug/L	1572	Standard
Sr	88	65.0	20.4				ug/L	73	Standard
C	12	60.0					mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	25.5	84.0				mg/L	6	Standard
Ho-1	165	20.0	66.1				mg/L	10	Standard
Er	166	23.3	65.5				mg/L	13	Standard
I	127	1926.8	28.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		111.783	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:02:12
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.184
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.386
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Mn	55	

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:02:12
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Wednesday, June 29, 2016 18:03:08
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84590.6	2.3				ug/L	74699	Standard
	Be	9	260.0	12.0	**0.2687**	0.034	12.7	ug/L	8	Standard
	Al	27	3603.8	43.5	**0.0082**	0.009	110.1	ug/L	1493	Standard
	Sc	45	65034.3	0.9				ug/L	61029	Standard
	Ti	47	43.7	32.9	**-0.0008**	0.035	4199.9	ug/L	35	Standard
	V	51	4694.6	2.2	**0.3894**	0.008	2.0	ug/L	1435	Standard
	Cr	52	14076.2	0.6	**0.7639**	0.008	1.0	ug/L	7635	Standard
	Cr	53	1435.1	6.7	**1.1315**	0.098	8.6	ug/L	363	Standard
	Mn	55	4379.0	3.5	**0.4438**	0.017	3.9	ug/L	747	Standard
	Co	59	3208.0	4.2	**0.4198**	0.015	3.5	ug/L	234	Standard
	Ni	60	3464.1	1.9	**1.6175**	0.036	2.2	ug/L	215	Standard
	Cu	65	1802.1	7.0	**0.8073**	0.055	6.8	ug/L	104	Standard
	Zn	66	8137.9	1.9	**6.5566**	0.074	1.1	ug/L	363	Standard
>	**Ge**	72	526101.8	0.9				ug/L	495142	Standard
	As	75	520.8	10.5	**0.4970**	0.042	8.4	ug/L	-50	Standard
	Se	82	68.5	3.0	**0.4490**	0.013	3.0	ug/L	14	Standard
	Se-1	77	97.0	14.4	**0.6266**	0.169	26.9	ug/L	42	Standard
>	**Ga**	71	35.0	42.9				mg/L	42	Standard
	Rb	85	53.3	5.4				ug/L	17	Standard
	Y	89	462050.3	3.1				ug/L	417098	Standard
>	**Rh**	103	11.7	24.7				ug/L	13	Standard
	Mo	98	178.6	39.2	**0.0548**	0.024	44.0	ug/L	11	Standard
	Ag	107	2989.6	2.3	**0.4006**	0.011	2.7	ug/L	68	Standard
	Cd	111	543.3	9.6	**0.2264**	0.024	10.6	mg/L	2	Standard
	Cd	114	1493.6	11.4	**0.2605**	0.031	12.1	ug/L	12	Standard
>	**In**	115	651766.1	0.6				ug/L	587441	Standard
	Sn	118	551.7	10.5	**-0.0096**	0.009	97.6	ug/L	497	Standard
	Sb	123	2328.3	1.4	**0.4453**	0.009	1.9	ug/L	78	Standard
	Ba	135	1600.4	3.6	**0.6889**	0.030	4.3	ug/L	24	Standard
	Ce	140	6.7	86.6				ug/L	25	Standard
>	**Tb**	159	1010895.5	0.8				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	860.7	17.5	**0.0936**	0.016	16.7	ug/L	4	Standard
	Tl	205	718.4	10.6	**0.0918**	0.008	8.5	ug/L	2	Standard
	Pb	206	1533.1	9.5	**0.1909**	0.020	10.6	ug/L	240	Standard
	Pb	207	1247.1	5.1	**0.1700**	0.007	4.3	ug/L	198	Standard
	Pb	208	5357.0	3.3	**0.1870**	0.005	2.6	ug/L	883	Standard
	U	238	3040.0	2.0	**0.3994**	0.014	3.4	ug/L	3	Standard
>	**Bi**	209	573346.3	2.5				ug/L	523984	Standard

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 18:05:25
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.2178**	31.546	173.2	mg/L	0	Standard
Mg	24	50.0	50.0	**-0.0044**	0.031	706.6	mg/L	37	Standard
K	39	35.0	24.7	**-0.0415**	0.018	44.0	mg/L	38	Standard
Ca	43	26.7	43.3	**-0.9446**	0.676	71.5	mg/L	10	Standard
Fe	54	105.0	32.5	**-0.0546**	0.028	51.3	mg/L	179	Standard
Fe	57	230.0	24.7	**-0.3395**	0.189	55.5	mg/L	335	Standard
Sc-1	45	65034.3	0.9				mg/L	61029	Standard
Cl	35	27747.4	0.7				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	1776.8	9.7				ug/L	1547	Standard
P	31	14345.1	1.0				ug/L	15286	Standard
S	34	2316.8	2.0				ug/L	1572	Standard
Sr	88	78.3	41.5				ug/L	73	Standard
C	12	70.0	28.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	6.0	100.4				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	1788.4	30.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	134.333		
Al	27			
Sc	45			
Ti	47			
V	51	97.347		
Cr	52	95.485		
Cr	53			
Mn	55	88.762		
Co	59	104.961		
Ni	60	101.093		
Cu	65	100.908		
Zn	66	104.906		
Ge	72		106.253	
As	75	124.241		
Se	82	112.254		
Se-1	77			
Ga	71			

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 18:05:25
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	100.142		
Cd	111	94.325		
Cd	114			
> In	115			110.950
Sn	118			
Sb	123	111.327		
Ba	135	91.848		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	116.967		
Tl	205			
Pb	206			
Pb	207			
Pb	208	93.500		
U	238	99.844		
> Bi	209			109.420
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 8	Be	9	

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 18:05:25
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBS M9 WG574192-02
Sample Date/Time: Wednesday, June 29, 2016 18:06:21
Number of Replicates: 3
Autosampler Position: 237
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84464.8	1.8				ug/L	74699	Standard
	Be	9	13.3	94.4	**0.0212**	0.013	60.4	ug/L	8	Standard
	Al	27	11484.4	1.6	**0.0557**	0.001	1.2	ug/L	1493	Standard
	Sc	45	65588.4	2.6				ug/L	61029	Standard
	Ti	47	51.3	20.8	**0.0150**	0.026	174.5	ug/L	35	Standard
	V	51	1499.5	1.3	**-0.0137**	0.003	18.5	ug/L	1435	Standard
	Cr	52	8909.3	1.2	**0.0773**	0.005	6.6	ug/L	7635	Standard
	Cr	53	853.4	5.0	**0.5155**	0.034	6.6	ug/L	363	Standard
	Mn	55	1324.4	4.8	**0.0262**	0.011	41.6	ug/L	747	Standard
	Co	59	237.0	9.4	**-0.0044**	0.004	80.7	ug/L	234	Standard
	Ni	60	317.3	8.0	**0.0315**	0.014	46.0	ug/L	215	Standard
	Cu	65	277.0	2.0	**0.0327**	0.005	14.4	ug/L	104	Standard
	Zn	66	2117.5	3.0	**1.3211**	0.055	4.2	ug/L	363	Standard
>	**Ge**	72	540634.9	1.6				ug/L	495142	Standard
	As	75	-19.3	191.5	**0.0441**	0.030	68.3	ug/L	-50	Standard
	Se	82	22.9	30.4	**0.0532**	0.060	113.4	ug/L	14	Standard
	Se-1	77	66.3	3.1	**0.2096**	0.028	13.1	ug/L	42	Standard
>	**Ga**	71	40.0	45.1				mg/L	42	Standard
	Rb	85	45.0	40.1				ug/L	17	Standard
	Y	89	495113.6	1.2				ug/L	417098	Standard
>	**Rh**	103	20.0	50.0				ug/L	13	Standard
	Mo	98	103.4	4.5	**0.0285**	0.002	7.3	ug/L	11	Standard
	Ag	107	105.3	22.1	**0.0013**	0.003	227.8	ug/L	68	Standard
	Cd	111	14.8	35.7	**-0.0017**	0.002	127.2	mg/L	2	Standard
	Cd	114	30.3	51.7	**0.0031**	0.003	86.4	ug/L	12	Standard
>	**In**	115	666604.9	1.7				ug/L	587441	Standard
	Sn	118	860.0	2.1	**0.0351**	0.004	11.7	ug/L	497	Standard
	Sb	123	400.3	27.4	**0.0716**	0.022	30.7	ug/L	78	Standard
	Ba	135	277.0	2.3	**0.1032**	0.001	0.8	ug/L	24	Standard
	Ce	140	75.0	20.0				ug/L	25	Standard
>	**Tb**	159	1002064.8	2.5				ug/L	939701	Standard
	Ho	165	13.3	21.7				ug/L	10	Standard
	Tl	203	23.7	14.8	**0.0009**	0.000	40.1	ug/L	4	Standard
	Tl	205	18.3	56.8	**0.0026**	0.001	51.0	ug/L	2	Standard
	Pb	206	383.3	7.2	**-0.0041**	0.004	93.5	ug/L	240	Standard
	Pb	207	304.7	4.5	**-0.0100**	0.002	23.2	ug/L	198	Standard
	Pb	208	1385.4	1.0	**-0.0050**	0.001	21.8	ug/L	883	Standard
	U	238	16.0	18.8	**-0.0023**	0.000	16.8	ug/L	3	Standard
>	**Bi**	209	560531.0	1.7				ug/L	523984	Standard

Sample ID: PBS M9 WG574192-02
Report Date/Time: Wednesday, June 29, 2016 18:08:38
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	51.7	11.2	**-0.0029**	0.008	286.8	mg/L	37	Standard
K	39	28.3	27.0	**-0.0560**	0.015	26.7	mg/L	38	Standard
Ca	43	25.0	34.6	**-1.0611**	0.483	45.5	mg/L	10	Standard
Fe	54	209.4	8.6	**0.0268**	0.012	44.3	mg/L	179	Standard
Fe	57	206.7	9.2	**-0.4211**	0.048	11.4	mg/L	335	Standard
> **Sc-1**	45	65588.4	2.6				mg/L	61029	Standard
Cl	35	29295.6	1.0				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	3250.3	3.7				ug/L	1547	Standard
P	31	19926.4	1.7				ug/L	15286	Standard
S	34	2431.9	7.4				ug/L	1572	Standard
Sr	88	63.3	32.9				ug/L	73	Standard
C	12	60.0	16.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	23.2	64.8				mg/L	6	Standard
Ho-1	165	13.3	21.7				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	1665.1	25.5				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		113.074	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		109.188	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBS M9 WG574192-02
Report Date/Time: Wednesday, June 29, 2016 18:08:38
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.476
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.975
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBS M9 WG574192-02
Report Date/Time: Wednesday, June 29, 2016 18:08:38
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSS M9 WG574192-03
Sample Date/Time: Wednesday, June 29, 2016 18:09:32
Number of Replicates: 3
Autosampler Position: 238
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83350.4	2.2				ug/L	74699	Standard
	Be	9	28323.1	1.8	**28.8492**	0.865	3.0	ug/L	8	Standard
	Al	27	10196.8	1.6	**0.0488**	0.002	4.6	ug/L	1493	Standard
	Sc	45	65183.2	0.6				ug/L	61029	Standard
	Ti	47	95.7	13.6	**0.1113**	0.027	24.3	ug/L	35	Standard
	V	51	224256.1	0.6	**25.7588**	0.522	2.0	ug/L	1435	Standard
	Cr	52	224696.6	1.0	**25.4792**	0.636	2.5	ug/L	7635	Standard
	Cr	53	27321.3	1.7	**25.5112**	0.743	2.9	ug/L	363	Standard
	Mn	55	202745.7	1.5	**25.3082**	0.750	3.0	ug/L	747	Standard
	Co	59	203564.8	2.3	**26.9126**	1.011	3.8	ug/L	234	Standard
	Ni	60	54379.7	0.7	**25.2634**	0.489	1.9	ug/L	215	Standard
	Cu	65	53975.6	1.0	**25.2465**	0.617	2.4	ug/L	104	Standard
	Zn	66	33582.8	1.1	**26.4199**	0.666	2.5	ug/L	363	Standard
>	**Ge**	72	566570.7	1.5				ug/L	495142	Standard
	As	75	31915.6	1.0	**24.9643**	0.450	1.8	ug/L	-50	Standard
	Se	82	3210.1	2.6	**25.4315**	0.960	3.8	ug/L	14	Standard
	Se-1	77	2232.8	4.1	**26.1060**	1.397	5.4	ug/L	42	Standard
>	**Ga**	71	91.7	13.7				mg/L	42	Standard
	Rb	85	96.7	31.2				ug/L	17	Standard
	Y	89	463441.3	2.1				ug/L	417098	Standard
>	**Rh**	103	25.0	40.0				ug/L	13	Standard
	Mo	98	76.3	5.4	**0.0207**	0.001	6.1	ug/L	11	Standard
	Ag	107	186996.6	1.0	**26.4732**	0.339	1.3	ug/L	68	Standard
	Cd	111	58916.9	1.6	**26.0048**	0.541	2.1	mg/L	2	Standard
	Cd	114	141808.0	1.8	**25.5238**	0.616	2.4	ug/L	12	Standard
>	**In**	115	636399.6	0.6				ug/L	587441	Standard
	Sn	118	665.0	9.8	**0.0103**	0.010	94.1	ug/L	497	Standard
	Sb	123	127977.1	1.2	**25.3000**	0.453	1.8	ug/L	78	Standard
	Ba	135	55490.8	0.9	**25.0111**	0.350	1.4	ug/L	24	Standard
	Ce	140	61.7	36.6				ug/L	25	Standard
>	**Tb**	159	977939.2	1.1				ug/L	939701	Standard
	Ho	165	5.0	100.0				ug/L	10	Standard
	Tl	203	232757.9	0.8	**26.2994**	0.658	2.5	ug/L	4	Standard
	Tl	205	213539.2	0.1	**27.8001**	0.582	2.1	ug/L	2	Standard
	Pb	206	150898.4	0.7	**26.2407**	0.649	2.5	ug/L	240	Standard
	Pb	207	129115.2	1.2	**25.2720**	0.800	3.2	ug/L	198	Standard
	Pb	208	547777.7	0.6	**27.1359**	0.615	2.3	ug/L	883	Standard
	U	238	197609.3	0.3	**26.7341**	0.481	1.8	ug/L	3	Standard
>	**Bi**	209	562718.9	2.0				ug/L	523984	Standard

Sample ID: LCSS M9 WG574192-03
Report Date/Time: Wednesday, June 29, 2016 18:11:49
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**35.9815**	31.158	86.6	mg/L	0	Standard
Mg	24	53.3	23.6	**-0.0007**	0.015	2138.9	mg/L	37	Standard
K	39	36.7	90.8	**-0.0380**	0.070	183.0	mg/L	38	Standard
Ca	43	26.7	28.6	**-0.9444**	0.462	48.9	mg/L	10	Standard
Fe	54	149.2	16.7	**-0.0198**	0.020	102.0	mg/L	179	Standard
Fe	57	258.3	10.7	**-0.2518**	0.092	36.6	mg/L	335	Standard
> **Sc-1**	45	65183.2	0.6				mg/L	61029	Standard
Cl	35	28068.6	2.1				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	1853.4	3.5				ug/L	1547	Standard
P	31	20221.8	1.7				ug/L	15286	Standard
S	34	2256.8	4.8				ug/L	1572	Standard
Sr	88	75.0	24.0				ug/L	73	Standard
C	12	113.3	53.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	2.9	203.0				mg/L	6	Standard
Ho-1	165	5.0	100.0				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	1621.8	26.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.582	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.426	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSS M9 WG574192-03
Report Date/Time: Wednesday, June 29, 2016 18:11:49
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.334
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.392
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSS M9 WG574192-03
Report Date/Time: Wednesday, June 29, 2016 18:11:49
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120601 WG574192-01
Sample Date/Time: Wednesday, June 29, 2016 18:12:44
Number of Replicates: 3
Autosampler Position: 239
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89463.4	1.2				ug/L	74699	Standard
	Be	9	386.7	13.6	**0.3740**	0.046	12.2	ug/L	8	Standard
	Al	27	7605562.2	1.6	**43.1976**	0.340	0.8	ug/L	1493	Standard
	Sc	45	72243.5	1.6				ug/L	61029	Standard
	Ti	47	9362.6	0.8	**21.8169**	0.679	3.1	ug/L	35	Standard
	V	51	123334.5	1.4	**14.2946**	0.165	1.2	ug/L	1435	Standard
	Cr	52	107034.1	1.6	**11.7966**	0.478	4.1	ug/L	7635	Standard
	Cr	53	12683.6	2.9	**11.8556**	0.642	5.4	ug/L	363	Standard
	Mn	55	2726083.5	2.1	**347.2409**	2.026	0.6	ug/L	747	Standard
	Co	59	29531.4	3.0	**3.9299**	0.043	1.1	ug/L	234	Standard
	Ni	60	23196.4	0.4	**10.8725**	0.312	2.9	ug/L	215	Standard
	Cu	65	77156.3	0.7	**36.6848**	0.692	1.9	ug/L	104	Standard
	Zn	66	748003.6	1.6	**607.1548**	7.539	1.2	ug/L	363	Standard
>	**Ge**	72	558122.4	2.5				ug/L	495142	Standard
	As	75	3312.8	2.4	**2.6835**	0.016	0.6	ug/L	-50	Standard
	Se	82	112.5	6.7	**0.7726**	0.083	10.8	ug/L	14	Standard
	Se-1	77	130.7	4.7	**0.9656**	0.087	9.0	ug/L	42	Standard
>	**Ga**	71	7793.7	3.0				mg/L	42	Standard
	Rb	85	97459.1	1.1				ug/L	17	Standard
	Y	89	585281.5	2.1				ug/L	417098	Standard
>	**Rh**	103	56.7	13.5				ug/L	13	Standard
	Mo	98	1826.8	4.5	**0.5887**	0.034	5.8	ug/L	11	Standard
	Ag	107	1189.0	7.2	**0.1453**	0.012	8.4	ug/L	68	Standard
	Cd	111	5739.0	1.3	**2.3714**	0.004	0.2	mg/L	2	Standard
	Cd	114	13674.5	3.2	**2.3090**	0.056	2.4	ug/L	12	Standard
>	**In**	115	677670.9	1.2				ug/L	587441	Standard
	Sn	118	2238.5	6.9	**0.2377**	0.027	11.3	ug/L	497	Standard
	Sb	123	1060.8	22.9	**0.1924**	0.043	22.3	ug/L	78	Standard
	Ba	135	275486.3	0.7	**116.6679**	1.279	1.1	ug/L	24	Standard
	Ce	140	521396.3	0.5				ug/L	25	Standard
>	**Tb**	159	1067665.8	1.0				ug/L	939701	Standard
	Ho	165	8012.1	4.7				ug/L	10	Standard
	Tl	203	658.7	5.9	**0.0698**	0.004	5.8	ug/L	4	Standard
	Tl	205	588.3	7.1	**0.0738**	0.005	7.2	ug/L	2	Standard
	Pb	206	280687.1	0.1	**46.9821**	0.294	0.6	ug/L	240	Standard
	Pb	207	217657.6	0.6	**40.9978**	0.453	1.1	ug/L	198	Standard
	Pb	208	937818.9	1.0	**44.7144**	0.652	1.5	ug/L	883	Standard
	U	238	5450.3	0.5	**0.7046**	0.003	0.4	ug/L	3	Standard
>	**Bi**	209	585148.7	0.6				ug/L	523984	Standard

Sample ID: L1606120601 WG574192-01
Report Date/Time: Wednesday, June 29, 2016 18:15:01
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	733.4	5.7	**0.7343**	0.035	4.8	mg/L	37	Standard
K	39	401.7	1.9	**0.6360**	0.019	3.0	mg/L	38	Standard
Ca	43	643.3	17.5	**31.7962**	5.845	18.4	mg/L	10	Standard
Fe	54	8818.1	2.4	**6.1777**	0.130	2.1	mg/L	179	Standard
Fe	57	2673.6	4.7	**6.5999**	0.243	3.7	mg/L	335	Standard
Sc-1	45	72243.5	1.6				mg/L	61029	Standard
Cl	35	29798.6	1.4				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	3707.1	5.8				ug/L	1547	Standard
P	31	20168.4	2.0				ug/L	15286	Standard
S	34	2373.5	6.9				ug/L	1572	Standard
Sr	88	256.7	13.8				ug/L	73	Standard
C	12	350.0	7.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	130.0	53.3				mg/L	3	Standard
Dy	164	13278.3	0.5				mg/L	6	Standard
Ho-1	165	8012.1	4.7				mg/L	10	Standard
Er	166	7608.6	3.8				mg/L	13	Standard
I	127	89948.0	2.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		119.765	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		112.720	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120601 WG574192-01
Report Date/Time: Wednesday, June 29, 2016 18:15:01
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	115.360
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	111.673
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120602
Sample Date/Time: Wednesday, June 29, 2016 18:15:55
Number of Replicates: 3
Autosampler Position: 240
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88784.4	1.1				ug/L	74699	Standard
	Be	9	281.7	5.7	**0.2767**	0.013	4.7	ug/L	8	Standard
	Al	27	7005404.7	2.4	**40.0898**	0.625	1.6	ug/L	1493	Standard
	Sc	45	70594.2	3.3				ug/L	61029	Standard
	Ti	47	10155.7	1.6	**24.3492**	0.468	1.9	ug/L	35	Standard
	V	51	153110.0	0.2	**18.3086**	0.101	0.6	ug/L	1435	Standard
	Cr	52	102957.6	0.7	**11.6568**	0.092	0.8	ug/L	7635	Standard
	Cr	53	11732.9	3.0	**11.2567**	0.305	2.7	ug/L	363	Standard
	Mn	55	2052735.8	1.1	**268.9604**	2.146	0.8	ug/L	747	Standard
	Co	59	33492.3	1.4	**4.5921**	0.057	1.2	ug/L	234	Standard
	Ni	60	26642.1	0.7	**12.8642**	0.053	0.4	ug/L	215	Standard
	Cu	65	40463.3	0.7	**19.7403**	0.183	0.9	ug/L	104	Standard
	Zn	66	608360.3	0.1	**507.8962**	1.599	0.3	ug/L	363	Standard
>	**Ge**	72	542468.6	0.4				ug/L	495142	Standard
	As	75	4231.0	2.7	**3.5076**	0.102	2.9	ug/L	-50	Standard
	Se	82	256.2	4.5	**1.9932**	0.099	5.0	ug/L	14	Standard
	Se-1	77	208.7	7.7	**1.9860**	0.210	10.6	ug/L	42	Standard
>	**Ga**	71	5872.8	1.9				mg/L	42	Standard
	Rb	85	68412.5	1.5				ug/L	17	Standard
	Y	89	570393.2	0.3				ug/L	417098	Standard
>	**Rh**	103	206.7	25.4				ug/L	13	Standard
	Mo	98	11374.8	0.4	**3.8637**	0.075	1.9	ug/L	11	Standard
	Ag	107	1353.7	2.3	**0.1754**	0.001	0.8	ug/L	68	Standard
	Cd	111	5491.2	1.2	**2.3732**	0.046	1.9	mg/L	2	Standard
	Cd	114	13400.9	1.9	**2.3668**	0.045	1.9	ug/L	12	Standard
>	**In**	115	648040.3	1.6				ug/L	587441	Standard
	Sn	118	1990.1	9.0	**0.2145**	0.033	15.4	ug/L	497	Standard
	Sb	123	589.0	13.2	**0.1100**	0.013	12.1	ug/L	78	Standard
	Ba	135	226861.9	1.0	**100.4795**	2.067	2.1	ug/L	24	Standard
	Ce	140	495305.9	0.7				ug/L	25	Standard
>	**Tb**	159	1028084.3	0.7				ug/L	939701	Standard
	Ho	165	7586.9	1.7				ug/L	10	Standard
	Tl	203	574.7	1.8	**0.0640**	0.002	2.9	ug/L	4	Standard
	Tl	205	501.7	7.3	**0.0663**	0.004	6.1	ug/L	2	Standard
	Pb	206	2490204.6	1.0	**439.6443**	9.561	2.2	ug/L	240	Standard
	Pb	207	1941338.7	1.0	**385.7578**	8.533	2.2	ug/L	198	Standard
	Pb	208	8788859.7	0.3	**442.0013**	4.418	1.0	ug/L	883	Standard
	U	238	27264.2	0.8	**3.7315**	0.055	1.5	ug/L	3	Standard
>	**Bi**	209	555601.8	1.3				ug/L	523984	Standard

Sample ID: L1606120602
Report Date/Time: Wednesday, June 29, 2016 18:18:12
Page 1

Approved: July 04, 2016

Na	23	15.0	66.7	**151.5495**	103.438	68.3	mg/L	0	Standard
Mg	24	875.0	7.4	**0.9112**	0.062	6.8	mg/L	37	Standard
K	39	308.3	17.3	**0.4774**	0.118	24.7	mg/L	38	Standard
Ca	43	2528.5	1.9	**135.7080**	7.063	5.2	mg/L	10	Standard
Fe	54	8423.7	4.4	**6.0457**	0.457	7.6	mg/L	179	Standard
Fe	57	3312.0	6.3	**8.6682**	0.763	8.8	mg/L	335	Standard
> **Sc-1**	45	70594.2	3.3				mg/L	61029	Standard
Cl	35	33743.5	0.4				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	4283.9	2.2				ug/L	1547	Standard
P	31	21890.8	1.3				ug/L	15286	Standard
S	34	2663.6	6.3				ug/L	1572	Standard
Sr	88	533.3	8.7				ug/L	73	Standard
C	12	600.0	5.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	100.0	26.5				mg/L	3	Standard
Dy	164	11590.5	1.3				mg/L	6	Standard
Ho-1	165	7586.9	1.7				mg/L	10	Standard
Er	166	6698.1	4.4				mg/L	13	Standard
I	127	108507.7	4.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		118.857	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.558	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120602
Report Date/Time: Wednesday, June 29, 2016 18:18:12
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	110.316
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	106.034
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120602
Report Date/Time: Wednesday, June 29, 2016 18:18:12
Page 3

Approved: July 04, 2016

Pb 206 Upper, S, EEE	Pb	206
Pb 207 Upper, S, EEE	Pb	207
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120603S WG574192-04
Sample Date/Time: Wednesday, June 29, 2016 18:19:07
Number of Replicates: 3
Autosampler Position: 241
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86561.6	0.8				ug/L	74699	Standard
	Be	9	28455.1	4.4	**27.8947**	1.031	3.7	ug/L	8	Standard
	Al	27	11326081.2	0.4	**66.4980**	0.812	1.2	ug/L	1493	Standard
	Sc	45	74033.8	1.5				ug/L	61029	Standard
	Ti	47	13830.6	1.3	**31.7218**	0.771	2.4	ug/L	35	Standard
	V	51	404083.1	0.7	**46.4723**	0.823	1.8	ug/L	1435	Standard
	Cr	52	396378.6	0.5	**45.6257**	0.448	1.0	ug/L	7635	Standard
	Cr	53	48917.2	0.9	**45.8335**	0.539	1.2	ug/L	363	Standard
	Mn	55	3610519.8	1.3	**452.1326**	4.314	1.0	ug/L	747	Standard
	Co	59	243070.7	1.0	**32.0670**	0.356	1.1	ug/L	234	Standard
	Ni	60	87954.6	1.3	**40.8489**	0.598	1.5	ug/L	215	Standard
	Cu	65	138890.7	1.2	**64.9805**	0.614	0.9	ug/L	104	Standard
	Zn	66	982893.1	0.9	**784.3980**	13.633	1.7	ug/L	363	Standard
>	**Ge**	72	567734.3	1.2				ug/L	495142	Standard
	As	75	35264.6	1.3	**27.5210**	0.578	2.1	ug/L	-50	Standard
	Se	82	3103.1	1.8	**24.5237**	0.601	2.4	ug/L	14	Standard
	Se-1	77	2151.2	2.8	**25.0663**	0.745	3.0	ug/L	42	Standard
>	**Ga**	71	14508.6	3.6				mg/L	42	Standard
	Rb	85	180073.9	1.5				ug/L	17	Standard
	Y	89	616764.0	1.9				ug/L	417098	Standard
>	**Rh**	103	85.0	11.8				ug/L	13	Standard
	Mo	98	3441.4	2.1	**1.1615**	0.033	2.8	ug/L	11	Standard
	Ag	107	170180.4	0.7	**23.5897**	0.078	0.3	ug/L	68	Standard
	Cd	111	63817.8	1.8	**27.5852**	0.736	2.7	mg/L	2	Standard
	Cd	114	155790.6	2.3	**27.4604**	0.877	3.2	ug/L	12	Standard
>	**In**	115	649912.7	0.9				ug/L	587441	Standard
	Sn	118	2755.3	3.8	**0.3317**	0.018	5.5	ug/L	497	Standard
	Sb	123	11515.0	1.7	**2.2251**	0.028	1.3	ug/L	78	Standard
	Ba	135	420339.6	0.5	**185.6246**	2.179	1.2	ug/L	24	Standard
	Ce	140	771001.7	1.2				ug/L	25	Standard
>	**Tb**	159	1033151.8	0.5				ug/L	939701	Standard
	Ho	165	11344.2	1.5				ug/L	10	Standard
	Tl	203	220759.0	0.5	**24.6088**	0.113	0.5	ug/L	4	Standard
	Tl	205	203485.6	0.4	**26.1371**	0.140	0.5	ug/L	2	Standard
	Pb	206	518067.3	0.6	**89.0511**	0.447	0.5	ug/L	240	Standard
	Pb	207	415032.4	0.7	**80.2896**	0.537	0.7	ug/L	198	Standard
	Pb	208	1813547.0	1.2	**88.8061**	1.200	1.4	ug/L	883	Standard
	U	238	196913.8	1.4	**26.2850**	0.405	1.5	ug/L	3	Standard
>	**Bi**	209	570191.4	0.2				ug/L	523984	Standard

Sample ID: L1606120603S WG574192-04
Report Date/Time: Wednesday, June 29, 2016 18:21:24
Page 1

Approved: July 04, 2016

Na	23	6.7	43.3	**63.0881**	26.284	41.7	mg/L	0	Standard
Mg	24	761.7	8.6	**0.7448**	0.057	7.6	mg/L	37	Standard
K	39	698.3	9.3	**1.1574**	0.103	8.9	mg/L	38	Standard
Ca	43	761.7	2.5	**37.1357**	0.619	1.7	mg/L	10	Standard
Fe	54	13661.1	2.3	**9.4129**	0.352	3.7	mg/L	179	Standard
Fe	57	3945.5	4.9	**9.9839**	0.617	6.2	mg/L	335	Standard
> **Sc-1**	45	74033.8	1.5				mg/L	61029	Standard
Cl	35	29672.4	3.3				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	5120.9	2.8				ug/L	1547	Standard
P	31	21184.8	2.3				ug/L	15286	Standard
S	34	2585.2	3.6				ug/L	1572	Standard
Sr	88	278.3	7.3				ug/L	73	Standard
C	12	453.3	23.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	180.0	28.9				mg/L	3	Standard
Dy	164	18295.9	3.4				mg/L	6	Standard
Ho-1	165	11344.2	1.5				mg/L	10	Standard
Er	166	10350.2	2.2				mg/L	13	Standard
I	127	155575.8	4.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		115.881	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.661	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120603S WG574192-04
Report Date/Time: Wednesday, June 29, 2016 18:21:24
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	110.635
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.818
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120603S WG574192-04
Report Date/Time: Wednesday, June 29, 2016 18:21:24
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120604SD WG574192-05
Sample Date/Time: Wednesday, June 29, 2016 18:22:18
Number of Replicates: 3
Autosampler Position: 242
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88318.4	2.5				ug/L	74699	Standard
	Be	9	28910.9	0.8	**27.7911**	0.552	2.0	ug/L	8	Standard
	Al	27	11678142.6	0.8	**67.2186**	1.350	2.0	ug/L	1493	Standard
	Sc	45	75929.8	1.2				ug/L	61029	Standard
	Ti	47	15202.6	0.9	**34.3685**	0.367	1.1	ug/L	35	Standard
	V	51	408247.6	0.9	**46.2682**	0.632	1.4	ug/L	1435	Standard
	Cr	52	411599.1	1.0	**46.7151**	0.586	1.3	ug/L	7635	Standard
	Cr	53	50226.5	2.4	**46.3783**	0.843	1.8	ug/L	363	Standard
	Mn	55	4064951.2	0.1	**501.7134**	5.151	1.0	ug/L	747	Standard
	Co	59	242783.8	0.7	**31.5662**	0.473	1.5	ug/L	234	Standard
	Ni	60	92537.7	0.8	**42.3583**	0.203	0.5	ug/L	215	Standard
	Cu	65	135365.1	1.4	**62.4070**	0.241	0.4	ug/L	104	Standard
	Zn	66	1025151.4	1.0	**806.2133**	6.015	0.7	ug/L	363	Standard
>	**Ge**	72	576071.4	1.0				ug/L	495142	Standard
	As	75	35892.9	0.1	**27.6040**	0.298	1.1	ug/L	-50	Standard
	Se	82	2975.5	0.8	**23.1647**	0.155	0.7	ug/L	14	Standard
	Se-1	77	2118.1	0.7	**24.3051**	0.149	0.6	ug/L	42	Standard
>	**Ga**	71	16053.5	1.5				mg/L	42	Standard
	Rb	85	192660.5	0.7				ug/L	17	Standard
	Y	89	633758.5	0.9				ug/L	417098	Standard
>	**Rh**	103	91.7	22.7				ug/L	13	Standard
	Mo	98	2104.9	2.7	**0.6926**	0.023	3.3	ug/L	11	Standard
	Ag	107	176068.8	0.4	**23.8725**	0.248	1.0	ug/L	68	Standard
	Cd	111	64701.8	0.7	**27.3512**	0.177	0.6	mg/L	2	Standard
	Cd	114	157399.0	1.1	**27.1310**	0.176	0.6	ug/L	12	Standard
>	**In**	115	664480.0	1.2				ug/L	587441	Standard
	Sn	118	2676.9	2.8	**0.3105**	0.015	4.8	ug/L	497	Standard
	Sb	123	10363.1	2.2	**1.9584**	0.055	2.8	ug/L	78	Standard
	Ba	135	439420.6	0.3	**189.8023**	2.264	1.2	ug/L	24	Standard
	Ce	140	842883.5	1.0				ug/L	25	Standard
>	**Tb**	159	1049787.1	1.3				ug/L	939701	Standard
	Ho	165	12211.6	1.7				ug/L	10	Standard
	Tl	203	222058.4	0.3	**24.4396**	0.316	1.3	ug/L	4	Standard
	Tl	205	202828.8	1.6	**25.7185**	0.095	0.4	ug/L	2	Standard
	Pb	206	574235.7	0.3	**97.4667**	1.839	1.9	ug/L	240	Standard
	Pb	207	457211.7	0.6	**87.3408**	1.861	2.1	ug/L	198	Standard
	Pb	208	1979847.3	0.4	**95.7233**	1.171	1.2	ug/L	883	Standard
	U	238	197274.4	0.9	**26.0018**	0.641	2.5	ug/L	3	Standard
>	**Bi**	209	577596.9	1.6				ug/L	523984	Standard

Sample ID: L1606120604SD WG574192-05
Report Date/Time: Wednesday, June 29, 2016 18:24:35
Page 1

Approved: July 04, 2016

Na	23	5.0	0.0	**46.3267**	0.533	1.2	mg/L	0	Standard
Mg	24	961.7	5.5	**0.9325**	0.047	5.0	mg/L	37	Standard
K	39	738.4	15.4	**1.1991**	0.216	18.0	mg/L	38	Standard
Ca	43	858.4	11.4	**41.0973**	5.443	13.2	mg/L	10	Standard
Fe	54	16143.5	2.3	**10.8618**	0.122	1.1	mg/L	179	Standard
Fe	57	4519.0	2.4	**11.2744**	0.427	3.8	mg/L	335	Standard
Sc-1	45	75929.8	1.2				mg/L	61029	Standard
Cl	35	29136.0	1.4				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	5384.3	8.1				ug/L	1547	Standard
P	31	20994.5	3.6				ug/L	15286	Standard
S	34	2526.9	1.1				ug/L	1572	Standard
Sr	88	265.0	8.2				ug/L	73	Standard
C	12	466.7	15.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	193.3	16.6				mg/L	3	Standard
Dy	164	19678.1	2.3				mg/L	6	Standard
Ho-1	165	12211.6	1.7				mg/L	10	Standard
Er	166	11454.3	2.8				mg/L	13	Standard
I	127	133210.9	5.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		118.233	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		116.345	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120604SD WG574192-05
Report Date/Time: Wednesday, June 29, 2016 18:24:35
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	113.114
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	110.232
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120604SD WG574192-05
Report Date/Time: Wednesday, June 29, 2016 18:24:35
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605
Sample Date/Time: Wednesday, June 29, 2016 18:25:30
Number of Replicates: 3
Autosampler Position: 243
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90694.1	2.2				ug/L	74699	Standard
	Be	9	526.7	6.5	**0.5004**	0.031	6.1	ug/L	8	Standard
	Al	27	8237088.6	1.8	**46.1607**	0.953	2.1	ug/L	1493	Standard
	Sc	45	76241.4	2.3				ug/L	61029	Standard
	Ti	47	9939.3	1.0	**23.0163**	0.554	2.4	ug/L	35	Standard
	V	51	146708.3	2.5	**16.9369**	0.672	4.0	ug/L	1435	Standard
	Cr	52	214286.2	1.5	**24.4749**	0.700	2.9	ug/L	7635	Standard
	Cr	53	25651.7	1.8	**24.1415**	0.427	1.8	ug/L	363	Standard
	Mn	55	2369893.5	2.4	**300.0787**	12.076	4.0	ug/L	747	Standard
	Co	59	42091.1	2.2	**5.5844**	0.210	3.8	ug/L	234	Standard
	Ni	60	29674.7	1.4	**13.8532**	0.382	2.8	ug/L	215	Standard
	Cu	65	80108.6	1.5	**37.8555**	1.188	3.1	ug/L	104	Standard
	Zn	66	402296.4	1.3	**324.3434**	9.485	2.9	ug/L	363	Standard
>	**Ge**	72	561626.1	1.7				ug/L	495142	Standard
	As	75	4613.8	0.6	**3.6915**	0.048	1.3	ug/L	-50	Standard
	Se	82	73.3	11.4	**0.4501**	0.057	12.7	ug/L	14	Standard
	Se-1	77	107.3	9.4	**0.6736**	0.125	18.5	ug/L	42	Standard
>	**Ga**	71	10595.4	4.5				mg/L	42	Standard
	Rb	85	117987.7	1.5				ug/L	17	Standard
	Y	89	650208.5	0.3				ug/L	417098	Standard
>	**Rh**	103	85.0	38.6				ug/L	13	Standard
	Mo	98	1074.4	2.1	**0.3404**	0.008	2.3	ug/L	11	Standard
	Ag	107	1521.7	2.8	**0.1876**	0.005	2.8	ug/L	68	Standard
	Cd	111	14943.9	2.9	**6.1278**	0.124	2.0	mg/L	2	Standard
	Cd	114	36335.4	1.3	**6.0807**	0.042	0.7	ug/L	12	Standard
>	**In**	115	684218.1	1.0				ug/L	587441	Standard
	Sn	118	4600.7	0.6	**0.5815**	0.007	1.3	ug/L	497	Standard
	Sb	123	346.8	6.2	**0.0596**	0.003	5.7	ug/L	78	Standard
	Ba	135	305952.3	0.8	**128.3272**	1.108	0.9	ug/L	24	Standard
	Ce	140	926252.0	0.8				ug/L	25	Standard
>	**Tb**	159	1067500.9	0.3				ug/L	939701	Standard
	Ho	165	12695.3	1.4				ug/L	10	Standard
	Tl	203	780.0	9.1	**0.0850**	0.009	10.5	ug/L	4	Standard
	Tl	205	700.0	0.7	**0.0898**	0.001	1.1	ug/L	2	Standard
	Pb	206	317389.9	0.7	**54.3657**	1.082	2.0	ug/L	240	Standard
	Pb	207	249270.6	1.1	**48.0507**	1.100	2.3	ug/L	198	Standard
	Pb	208	1104553.9	2.1	**53.8955**	1.488	2.8	ug/L	883	Standard
	U	238	5524.0	2.1	**0.7309**	0.022	3.0	ug/L	3	Standard
>	**Bi**	209	572007.3	1.4				ug/L	523984	Standard

Sample ID: L1606120605
Report Date/Time: Wednesday, June 29, 2016 18:27:47
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**77.0658**	27.245	35.4	mg/L	0	Standard
Mg	24	545.0	4.9	**0.4986**	0.038	7.7	mg/L	37	Standard
K	39	528.3	15.8	**0.8195**	0.134	16.4	mg/L	38	Standard
Ca	43	966.7	6.0	**46.3930**	3.592	7.7	mg/L	10	Standard
Fe	54	12851.8	4.3	**8.5861**	0.399	4.6	mg/L	179	Standard
Fe	57	3822.2	7.3	**9.3402**	1.007	10.8	mg/L	335	Standard
Sc-1	45	76241.4	2.3				mg/L	61029	Standard
Cl	35	29717.8	2.2				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	3207.0	9.1				ug/L	1547	Standard
P	31	20467.1	0.8				ug/L	15286	Standard
S	34	2505.2	3.0				ug/L	1572	Standard
Sr	88	263.3	8.6				ug/L	73	Standard
C	12	196.7	5.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	123.3	46.8				mg/L	3	Standard
Dy	164	19864.2	0.8				mg/L	6	Standard
Ho-1	165	12695.3	1.4				mg/L	10	Standard
Er	166	11821.3	4.6				mg/L	13	Standard
I	127	41200.9	0.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		121.413	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.427	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120605
Report Date/Time: Wednesday, June 29, 2016 18:27:47
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	116.474
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	109.165
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606120605
Report Date/Time: Wednesday, June 29, 2016 18:27:47
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605PS WG574690-01
Sample Date/Time: Wednesday, June 29, 2016 18:28:41
Number of Replicates: 3
Autosampler Position: 244
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85371.5	1.3				ug/L	74699	Standard
	Be	9	57838.5	0.7	**57.4989**	0.995	1.7	ug/L	8	Standard
	Al	27	8190235.3	0.9	**48.7593**	1.052	2.2	ug/L	1493	Standard
	Sc	45	74174.5	1.1				ug/L	61029	Standard
	Ti	47	9905.9	1.1	**22.7993**	0.140	0.6	ug/L	35	Standard
	V	51	615589.0	1.1	**71.2575**	1.604	2.3	ug/L	1435	Standard
	Cr	52	662720.1	2.0	**77.3523**	0.990	1.3	ug/L	7635	Standard
	Cr	53	81212.1	1.6	**76.6855**	0.228	0.3	ug/L	363	Standard
	Mn	55	2850355.8	1.2	**358.7372**	7.553	2.1	ug/L	747	Standard
	Co	59	453959.9	1.8	**60.2304**	1.747	2.9	ug/L	234	Standard
	Ni	60	139316.4	1.9	**65.0924**	0.607	0.9	ug/L	215	Standard
	Cu	65	189057.7	1.5	**88.9267**	0.405	0.5	ug/L	104	Standard
	Zn	66	476114.4	1.0	**381.6097**	5.307	1.4	ug/L	363	Standard
>	**Ge**	72	564919.1	1.3				ug/L	495142	Standard
	As	75	71672.0	0.5	**56.1469**	0.522	0.9	ug/L	-50	Standard
	Se	82	6681.0	0.4	**53.2203**	0.482	0.9	ug/L	14	Standard
	Se-1	77	4575.0	1.8	**54.2855**	1.161	2.1	ug/L	42	Standard
>	**Ga**	71	10925.6	2.3				mg/L	42	Standard
	Rb	85	119128.2	1.4				ug/L	17	Standard
	Y	89	658944.3	2.5				ug/L	417098	Standard
>	**Rh**	103	98.3	24.0				ug/L	13	Standard
	Mo	98	1073.3	2.3	**0.3534**	0.015	4.1	ug/L	11	Standard
	Ag	107	364831.9	0.7	**49.9049**	1.416	2.8	ug/L	68	Standard
	Cd	111	132655.1	1.4	**56.5569**	1.029	1.8	mg/L	2	Standard
	Cd	114	321468.3	0.7	**55.8946**	1.680	3.0	ug/L	12	Standard
>	**In**	115	659074.5	2.5				ug/L	587441	Standard
	Sn	118	1318.4	1.7	**0.1066**	0.008	7.5	ug/L	497	Standard
	Sb	123	258576.1	0.4	**49.3799**	1.200	2.4	ug/L	78	Standard
	Ba	135	409891.1	0.5	**178.5483**	4.068	2.3	ug/L	24	Standard
	Ce	140	929186.0	0.4				ug/L	25	Standard
>	**Tb**	159	1023915.1	1.8				ug/L	939701	Standard
	Ho	165	12935.5	3.1				ug/L	10	Standard
	Tl	203	455240.8	0.8	**52.4471**	0.837	1.6	ug/L	4	Standard
	Tl	205	420642.4	0.2	**55.8346**	0.392	0.7	ug/L	2	Standard
	Pb	206	614831.5	0.9	**109.2366**	1.843	1.7	ug/L	240	Standard
	Pb	207	503849.3	0.9	**100.7500**	1.617	1.6	ug/L	198	Standard
	Pb	208	2170185.0	1.2	**109.8441**	2.227	2.0	ug/L	883	Standard
	U	238	393858.5	1.0	**54.3381**	1.023	1.9	ug/L	3	Standard
>	**Bi**	209	551780.7	0.8				ug/L	523984	Standard

Sample ID: L1606120605PS WG574690-01
Report Date/Time: Wednesday, June 29, 2016 18:30:58
Page 1

Approved: July 04, 2016

Na	23	10.0	86.6	**95.2332**	83.153	87.3	mg/L	0	Standard
Mg	24	453.3	5.6	**0.4164**	0.027	6.4	mg/L	37	Standard
K	39	480.0	7.5	**0.7589**	0.067	8.8	mg/L	38	Standard
Ca	43	940.0	10.6	**46.3668**	5.650	12.2	mg/L	10	Standard
Fe	54	12349.5	1.2	**8.4763**	0.023	0.3	mg/L	179	Standard
Fe	57	3702.1	2.7	**9.2777**	0.167	1.8	mg/L	335	Standard
Sc-1	45	74174.5	1.1				mg/L	61029	Standard
Cl	35	30087.8	0.9				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	3260.4	2.1				ug/L	1547	Standard
P	31	20787.6	1.7				ug/L	15286	Standard
S	34	2500.2	2.8				ug/L	1572	Standard
Sr	88	245.0	12.4				ug/L	73	Standard
C	12	160.0	54.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	103.3	22.3				mg/L	3	Standard
Dy	164	20338.3	3.5				mg/L	6	Standard
Ho-1	165	12935.5	3.1				mg/L	10	Standard
Er	166	11817.9	3.9				mg/L	13	Standard
I	127	39775.5	3.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		114.288	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.092	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120605PS WG574690-01
Report Date/Time: Wednesday, June 29, 2016 18:30:58
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	112.194
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	105.305
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120605PS WG574690-01
Report Date/Time: Wednesday, June 29, 2016 18:30:58
Page 3

Approved: July 04, 2016

Pb 206 Upper, S, EEE	Pb	206
Pb 207 Upper, S, EEE	Pb	207
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605SDL WG574690-02
Sample Date/Time: Wednesday, June 29, 2016 18:31:53
Number of Replicates: 3
Autosampler Position: 245
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84476.5	1.1				ug/L	74699	Standard
	Be	9	178.3	18.7	**0.1869**	0.035	18.6	ug/L	8	Standard
	Al	27	1736036.7	3.1	**10.4319**	0.283	2.7	ug/L	1493	Standard
	Sc	45	72059.3	1.7				ug/L	61029	Standard
	Ti	47	2123.5	6.6	**4.6249**	0.308	6.7	ug/L	35	Standard
	V	51	32734.7	3.8	**3.4666**	0.135	3.9	ug/L	1435	Standard
	Cr	52	51911.1	1.7	**4.8949**	0.101	2.1	ug/L	7635	Standard
	Cr	53	5774.4	3.1	**4.9492**	0.156	3.2	ug/L	363	Standard
	Mn	55	492914.2	2.0	**59.6732**	1.219	2.0	ug/L	747	Standard
	Co	59	9193.8	6.0	**1.1391**	0.070	6.1	ug/L	234	Standard
	Ni	60	6665.5	3.0	**2.8840**	0.083	2.9	ug/L	215	Standard
	Cu	65	17774.1	0.7	**7.9654**	0.025	0.3	ug/L	104	Standard
	Zn	66	87339.7	0.8	**67.1049**	0.286	0.4	ug/L	363	Standard
>	**Ge**	72	585998.9	0.5				ug/L	495142	Standard
	As	75	970.2	5.1	**0.7914**	0.035	4.4	ug/L	-50	Standard
	Se	82	25.1	35.6	**0.0547**	0.068	124.2	ug/L	14	Standard
	Se-1	77	82.0	5.6	**0.3262**	0.050	15.4	ug/L	42	Standard
>	**Ga**	71	2373.5	5.4				mg/L	42	Standard
	Rb	85	23598.4	1.4				ug/L	17	Standard
	Y	89	515408.4	0.8				ug/L	417098	Standard
>	**Rh**	103	21.7	35.3				ug/L	13	Standard
	Mo	98	369.9	37.1	**0.1174**	0.046	39.0	ug/L	11	Standard
	Ag	107	699.0	38.5	**0.0822**	0.037	44.7	ug/L	68	Standard
	Cd	111	3080.1	3.0	**1.2985**	0.040	3.1	mg/L	2	Standard
	Cd	114	7389.1	5.7	**1.2756**	0.072	5.6	ug/L	12	Standard
>	**In**	115	662403.8	0.2				ug/L	587441	Standard
	Sn	118	660.0	13.5	**0.0055**	0.013	245.5	ug/L	497	Standard
	Sb	123	2141.2	12.8	**0.4025**	0.052	13.0	ug/L	78	Standard
	Ba	135	63716.0	0.4	**27.5914**	0.127	0.5	ug/L	24	Standard
	Ce	140	193219.3	1.3				ug/L	25	Standard
>	**Tb**	159	1027182.9	1.4				ug/L	939701	Standard
	Ho	165	2611.9	6.2				ug/L	10	Standard
	Tl	203	353.7	43.7	**0.0374**	0.017	46.4	ug/L	4	Standard
	Tl	205	421.7	67.1	**0.0541**	0.036	67.5	ug/L	2	Standard
	Pb	206	67035.7	2.8	**11.3788**	0.381	3.4	ug/L	240	Standard
	Pb	207	53097.6	2.6	**10.1377**	0.324	3.2	ug/L	198	Standard
	Pb	208	229986.8	1.7	**11.1165**	0.256	2.3	ug/L	883	Standard
	U	238	1351.4	18.8	**0.1748**	0.035	19.9	ug/L	3	Standard
>	**Bi**	209	574350.4	0.6				ug/L	523984	Standard

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:34:10
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	130.0	13.3	**0.0772**	0.020	26.0	mg/L	37	Standard
K	39	128.3	6.0	**0.1263**	0.018	14.2	mg/L	38	Standard
Ca	43	216.7	23.7	**9.0911**	2.889	31.8	mg/L	10	Standard
Fe	54	2661.6	0.9	**1.7733**	0.039	2.2	mg/L	179	Standard
Fe	57	965.0	5.3	**1.7027**	0.101	5.9	mg/L	335	Standard
> **Sc-1**	45	72059.3	1.7				mg/L	61029	Standard
Cl	35	30260.9	1.5				ug/L	20195	Standard
Kr	83	0.3	173.2				ug/L	4	Standard
Br	81	2210.2	10.2				ug/L	1547	Standard
P	31	20780.9	2.4				ug/L	15286	Standard
S	34	2430.2	2.3				ug/L	1572	Standard
Sr	88	83.3	13.9				ug/L	73	Standard
C	12	56.7	40.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	33.3	17.3				mg/L	3	Standard
Dy	164	4484.9	5.6				mg/L	6	Standard
Ho-1	165	2611.9	6.2				mg/L	10	Standard
Er	166	2556.9	13.6				mg/L	13	Standard
I	127	10483.6	6.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.090	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		118.350	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:34:10
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.761
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.612
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605SDL WG574690-02
Sample Date/Time: Wednesday, June 29, 2016 18:35:04
Number of Replicates: 3
Autosampler Position: 246
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83239.9	3.6				ug/L	74699	Standard
	Be	9	38.3	32.8	**0.0467**	0.013	27.9	ug/L	8	Standard
	Al	27	353337.7	1.0	**2.1460**	0.078	3.6	ug/L	1493	Standard
	Sc	45	70061.7	2.7				ug/L	61029	Standard
	Ti	47	444.3	5.4	**0.9019**	0.065	7.2	ug/L	35	Standard
	V	51	7638.3	1.2	**0.6761**	0.009	1.3	ug/L	1435	Standard
	Cr	52	17382.6	1.1	**0.9981**	0.034	3.5	ug/L	7635	Standard
	Cr	53	1533.4	2.6	**1.0999**	0.009	0.8	ug/L	363	Standard
	Mn	55	100723.4	1.5	**12.3208**	0.101	0.8	ug/L	747	Standard
	Co	59	1988.5	0.8	**0.2223**	0.008	3.5	ug/L	234	Standard
	Ni	60	1493.7	0.8	**0.5638**	0.012	2.1	ug/L	215	Standard
	Cu	65	3674.4	3.3	**1.5981**	0.060	3.8	ug/L	104	Standard
	Zn	66	19238.9	1.8	**14.7255**	0.173	1.2	ug/L	363	Standard
>	**Ge**	72	574541.2	2.2				ug/L	495142	Standard
	As	75	156.2	2.1	**0.1799**	0.005	2.7	ug/L	-50	Standard
	Se	82	23.9	29.5	**0.0491**	0.053	108.8	ug/L	14	Standard
	Se-1	77	58.0	21.2	**0.0603**	0.130	215.1	ug/L	42	Standard
>	**Ga**	71	518.3	4.3				mg/L	42	Standard
	Rb	85	4807.4	2.9				ug/L	17	Standard
	Y	89	501442.7	1.9				ug/L	417098	Standard
>	**Rh**	103	25.0	20.0				ug/L	13	Standard
	Mo	98	84.6	15.3	**0.0217**	0.004	18.8	ug/L	11	Standard
	Ag	107	254.3	9.6	**0.0208**	0.003	15.4	ug/L	68	Standard
	Cd	111	620.8	2.0	**0.2487**	0.004	1.5	mg/L	2	Standard
	Cd	114	1552.7	9.2	**0.2596**	0.025	9.8	ug/L	12	Standard
>	**In**	115	679737.8	0.5				ug/L	587441	Standard
	Sn	118	525.0	13.7	**-0.0171**	0.010	60.3	ug/L	497	Standard
	Sb	123	593.1	26.9	**0.1056**	0.030	28.3	ug/L	78	Standard
	Ba	135	12802.7	1.3	**5.3901**	0.099	1.8	ug/L	24	Standard
	Ce	140	39412.9	3.5				ug/L	25	Standard
>	**Tb**	159	1032176.9	1.0				ug/L	939701	Standard
	Ho	165	550.0	19.7				ug/L	10	Standard
	Tl	203	66.7	15.4	**0.0056**	0.001	20.8	ug/L	4	Standard
	Tl	205	78.3	41.0	**0.0102**	0.004	40.5	ug/L	2	Standard
	Pb	206	13522.4	0.6	**2.2309**	0.023	1.0	ug/L	240	Standard
	Pb	207	10801.2	2.0	**1.9994**	0.022	1.1	ug/L	198	Standard
	Pb	208	47468.1	0.8	**2.2283**	0.042	1.9	ug/L	883	Standard
	U	238	258.3	8.4	**0.0297**	0.002	8.2	ug/L	3	Standard
>	**Bi**	209	576211.7	1.3				ug/L	523984	Standard

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:37:21
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	16.4	**-0.0125**	0.009	74.1	mg/L	37	Standard
K	39	58.3	35.7	**-0.0022**	0.039	1787.9	mg/L	38	Standard
Ca	43	95.0	32.0	**2.7333**	1.777	65.0	mg/L	10	Standard
Fe	54	697.3	2.7	**0.3769**	0.020	5.3	mg/L	179	Standard
Fe	57	380.0	7.9	**0.0519**	0.104	201.0	mg/L	335	Standard
> **Sc-1**	45	70061.7	2.7				mg/L	61029	Standard
Cl	35	30787.9	1.4				ug/L	20195	Standard
Kr	83	2.3	107.9				ug/L	4	Standard
Br	81	2130.1	1.6				ug/L	1547	Standard
P	31	20966.2	4.2				ug/L	15286	Standard
S	34	2485.2	3.8				ug/L	1572	Standard
Sr	88	90.0	33.3				ug/L	73	Standard
C	12	40.0	66.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	3	Standard
Dy	164	970.4	12.1				mg/L	6	Standard
Ho-1	165	550.0	19.7				mg/L	10	Standard
Er	166	480.0	14.6				mg/L	13	Standard
I	127	3620.4	13.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.434	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		116.036	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:37:21
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		115.712
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.967
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 18:38:18
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	82768.9	2.2				ug/L	74699	Standard
	Be	9	55586.8	2.9	**56.9851**	0.437	0.8	ug/L	8	Standard
	Al	27	8619622.5	1.7	**52.9268**	0.666	1.3	ug/L	1493	Standard
	Sc	45	69535.9	1.6				ug/L	61029	Standard
	Ti	47	45959.0	2.5	**108.3957**	2.700	2.5	ug/L	35	Standard
	V	51	452339.2	1.9	**53.3985**	1.095	2.1	ug/L	1435	Standard
	Cr	52	452917.5	1.6	**53.6621**	0.813	1.5	ug/L	7635	Standard
	Cr	53	54972.9	2.4	**52.8966**	1.385	2.6	ug/L	363	Standard
	Mn	55	414134.0	1.2	**53.0785**	0.725	1.4	ug/L	747	Standard
	Co	59	411024.9	0.5	**55.6590**	0.257	0.5	ug/L	234	Standard
	Ni	60	111206.0	1.9	**53.0242**	0.984	1.9	ug/L	215	Standard
	Cu	65	110023.1	1.1	**52.7940**	0.716	1.4	ug/L	104	Standard
	Zn	66	64745.4	0.8	**52.5833**	0.259	0.5	ug/L	363	Standard
>	Ge	72	553341.4	0.3				ug/L	495142	Standard
	As	75	65551.1	0.4	**52.4265**	0.227	0.4	ug/L	-50	Standard
	Se	82	6664.4	1.0	**54.1963**	0.405	0.7	ug/L	14	Standard
	Se-1	77	4501.0	1.7	**54.5222**	0.959	1.8	ug/L	42	Standard
>	Ga	71	63.3	38.9				mg/L	42	Standard
	Rb	85	948.4	4.5				ug/L	17	Standard
	Y	89	490323.1	1.2				ug/L	417098	Standard
>	Rh	103	18.3	41.7				ug/L	13	Standard
	Mo	98	304786.5	0.3	**99.6729**	0.315	0.3	ug/L	11	Standard
	Ag	107	371315.6	1.3	**49.6520**	0.733	1.5	ug/L	68	Standard
	Cd	111	119443.2	1.0	**49.7891**	0.404	0.8	mg/L	2	Standard
	Cd	114	290803.0	0.5	**49.4253**	0.213	0.4	ug/L	12	Standard
>	In	115	673907.8	0.2				ug/L	587441	Standard
	Sn	118	339134.4	2.1	**50.5123**	1.079	2.1	ug/L	497	Standard
	Sb	123	262026.6	0.8	**48.9184**	0.455	0.9	ug/L	78	Standard
	Ba	135	111822.5	0.7	**47.6088**	0.434	0.9	ug/L	24	Standard
	Ce	140	206.7	14.6				ug/L	25	Standard
>	Tb	159	1024624.4	0.5				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	457896.5	0.3	**51.5737**	0.953	1.8	ug/L	4	Standard
	Tl	205	420019.2	1.0	**54.5041**	0.795	1.5	ug/L	2	Standard
	Pb	206	290158.6	1.2	**50.3552**	0.541	1.1	ug/L	240	Standard
	Pb	207	256755.7	1.2	**50.1530**	0.630	1.3	ug/L	198	Standard
	Pb	208	1071548.3	0.3	**52.9883**	1.261	2.4	ug/L	883	Standard
	U	238	389229.1	0.1	**52.5000**	1.135	2.2	ug/L	3	Standard
>	Bi	209	564495.9	2.1				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 18:40:35
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.7614**	29.023	173.2	mg/L	0	Standard
Mg	24	4642.4	0.2	**5.1955**	0.100	1.9	mg/L	37	Standard
K	39	2168.5	2.0	**4.0939**	0.144	3.5	mg/L	38	Standard
Ca	43	86.7	35.3	**2.2659**	1.624	71.7	mg/L	10	Standard
Fe	54	7113.4	2.2	**5.1545**	0.030	0.6	mg/L	179	Standard
Fe	57	2041.8	2.3	**5.0193**	0.241	4.8	mg/L	335	Standard
> **Sc-1**	45	69535.9	1.6				mg/L	61029	Standard
Cl	35	30041.8	1.0				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	1843.4	4.1				ug/L	1547	Standard
P	31	21683.8	2.5				ug/L	15286	Standard
S	34	2823.6	5.0				ug/L	1572	Standard
Sr	88	53.3	39.0				ug/L	73	Standard
C	12	53.3	78.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	12.9	179.7				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	2093.5	24.6				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	113.970		
	Al	27	105.854		
	Sc	45			
	Ti	47	108.396		
	V	51	106.797		
	Cr	52	107.324		
	Cr	53			
	Mn	55	106.157		
	Co	59	111.318		
	Ni	60	106.048		
	Cu	65	105.588		
	Zn	66	105.167		
>	Ge	72		111.754	
	As	75	104.853		
	Se	82	108.393		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 18:40:35
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	99.673		
Ag	107	99.304		
Cd	111	99.578		
Cd	114			
In	115		114.719	
Sn	118	101.025		
Sb	123	97.837		
Ba	135	95.218		
Ce	140			
Tb	159			
Ho	165			
Tl	203	103.147		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.977		
U	238	105.000		
Bi	209		107.731	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 18:40:35
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 18:41:29
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83048.8	2.8				ug/L	74699	Standard
	Be	9	23.3	68.9	**0.0313**	0.016	52.2	ug/L	8	Standard
	Al	27	3078.7	25.1	**0.0054**	0.004	78.6	ug/L	1493	Standard
	Sc	45	71367.7	1.6				ug/L	61029	Standard
	Ti	47	35.3	33.5	**-0.0277**	0.025	91.3	ug/L	35	Standard
	V	51	1419.9	5.2	**-0.0309**	0.006	18.5	ug/L	1435	Standard
	Cr	52	7844.0	0.2	**-0.0958**	0.021	21.7	ug/L	7635	Standard
	Cr	53	443.3	5.7	**0.0910**	0.034	37.3	ug/L	363	Standard
	Mn	55	990.7	12.0	**-0.0236**	0.012	50.6	ug/L	747	Standard
	Co	59	281.0	9.0	**-0.0000**	0.003	39793.0	ug/L	234	Standard
	Ni	60	196.7	7.3	**-0.0319**	0.005	14.4	ug/L	215	Standard
	Cu	65	145.3	12.4	**-0.0353**	0.007	21.2	ug/L	104	Standard
	Zn	66	464.0	9.4	**-0.0826**	0.026	32.0	ug/L	363	Standard
>	**Ge**	72	564975.6	2.4				ug/L	495142	Standard
	As	75	-39.3	63.0	**0.0286**	0.020	68.3	ug/L	-50	Standard
	Se	82	26.6	20.4	**0.0737**	0.038	51.8	ug/L	14	Standard
	Se-1	77	58.7	15.8	**0.0805**	0.100	124.7	ug/L	42	Standard
>	**Ga**	71	41.7	25.0				mg/L	42	Standard
	Rb	85	35.0	75.6				ug/L	17	Standard
	Y	89	486175.7	0.9				ug/L	417098	Standard
>	**Rh**	103	21.7	58.1				ug/L	13	Standard
	Mo	98	275.2	23.2	**0.0821**	0.020	24.6	ug/L	11	Standard
	Ag	107	144.3	16.7	**0.0059**	0.003	51.4	ug/L	68	Standard
	Cd	111	18.2	29.8	**-0.0005**	0.002	416.8	mg/L	2	Standard
	Cd	114	40.0	61.5	**0.0045**	0.004	89.7	ug/L	12	Standard
>	**In**	115	690622.1	0.6				ug/L	587441	Standard
	Sn	118	806.7	10.4	**0.0228**	0.013	56.6	ug/L	497	Standard
	Sb	123	553.4	29.9	**0.0967**	0.030	31.3	ug/L	78	Standard
	Ba	135	70.0	42.9	**0.0130**	0.012	94.6	ug/L	24	Standard
	Ce	140	61.7	54.0				ug/L	25	Standard
>	**Tb**	159	1047202.6	1.6				ug/L	939701	Standard
	Ho	165	16.7	34.6				ug/L	10	Standard
	Tl	203	75.3	48.5	**0.0063**	0.004	58.9	ug/L	4	Standard
	Tl	205	73.3	27.6	**0.0093**	0.002	26.3	ug/L	2	Standard
	Pb	206	333.0	14.3	**-0.0156**	0.007	42.7	ug/L	240	Standard
	Pb	207	293.0	14.2	**-0.0150**	0.006	43.1	ug/L	198	Standard
	Pb	208	1234.7	14.5	**-0.0154**	0.007	47.2	ug/L	883	Standard
	U	238	61.3	33.0	**0.0035**	0.002	69.2	ug/L	3	Standard
>	**Bi**	209	587080.9	2.4				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:43:46
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	38.3	15.1	**-0.0227**	0.007	30.4	mg/L	37	Standard
K	39	33.3	22.9	**-0.0512**	0.014	27.0	mg/L	38	Standard
Ca	43	28.3	50.9	**-0.9868**	0.802	81.3	mg/L	10	Standard
Fe	54	132.7	11.9	**-0.0421**	0.011	25.9	mg/L	179	Standard
Fe	57	280.0	10.7	**-0.2593**	0.100	38.4	mg/L	335	Standard
Sc-1	45	71367.7	1.6				mg/L	61029	Standard
Cl	35	29768.6	1.4				ug/L	20195	Standard
Kr	83	1.3	173.2				ug/L	4	Standard
Br	81	2106.8	7.7				ug/L	1547	Standard
P	31	21775.6	2.2				ug/L	15286	Standard
S	34	2523.5	3.0				ug/L	1572	Standard
Sr	88	70.0	49.5				ug/L	73	Standard
C	12	66.7	48.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	19.2	48.3				mg/L	6	Standard
Ho-1	165	16.7	34.6				mg/L	10	Standard
Er	166	16.7	91.7				mg/L	13	Standard
I	127	1905.1	33.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.104	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:43:46
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.564
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		112.042
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:43:46
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120701
Sample Date/Time: Wednesday, June 29, 2016 18:44:42
Number of Replicates: 3
Autosampler Position: 247
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82529.2	0.7				ug/L	74699	Standard
	Be	9	748.4	11.5	**0.7773**	0.092	11.9	ug/L	8	Standard
	Al	27	5186124.5	1.1	**31.9282**	0.368	1.2	ug/L	1493	Standard
	Sc	45	73214.8	0.3				ug/L	61029	Standard
	Ti	47	12465.8	1.5	**28.7779**	0.271	0.9	ug/L	35	Standard
	V	51	254152.3	1.7	**29.3595**	0.344	1.2	ug/L	1435	Standard
	Cr	52	145419.2	1.0	**16.2106**	0.127	0.8	ug/L	7635	Standard
	Cr	53	17483.4	2.8	**16.2818**	0.208	1.3	ug/L	363	Standard
	Mn	55	3993207.5	1.0	**503.6001**	3.344	0.7	ug/L	747	Standard
	Co	59	68748.0	1.0	**9.1067**	0.083	0.9	ug/L	234	Standard
	Ni	60	27938.1	3.3	**12.9787**	0.214	1.6	ug/L	215	Standard
	Cu	65	24664.7	1.5	**11.5352**	0.050	0.4	ug/L	104	Standard
	Zn	66	148101.8	1.7	**118.6204**	0.288	0.2	ug/L	363	Standard
>	**Ge**	72	563786.2	1.7				ug/L	495142	Standard
	As	75	6453.5	1.8	**5.1196**	0.031	0.6	ug/L	-50	Standard
	Se	82	62.5	2.1	**0.3624**	0.012	3.2	ug/L	14	Standard
	Se-1	77	114.0	16.7	**0.7504**	0.247	32.9	ug/L	42	Standard
>	**Ga**	71	11409.3	2.4				mg/L	42	Standard
	Rb	85	132848.2	0.6				ug/L	17	Standard
	Y	89	719521.7	0.4				ug/L	417098	Standard
>	**Rh**	103	43.3	43.7				ug/L	13	Standard
	Mo	98	809.7	3.4	**0.2640**	0.011	4.2	ug/L	11	Standard
	Ag	107	5298.9	1.5	**0.7088**	0.004	0.5	ug/L	68	Standard
	Cd	111	3980.4	2.6	**1.6822**	0.055	3.3	mg/L	2	Standard
	Cd	114	9713.7	1.4	**1.6793**	0.030	1.8	ug/L	12	Standard
>	**In**	115	661793.4	1.1				ug/L	587441	Standard
	Sn	118	911.7	2.8	**0.0439**	0.005	10.9	ug/L	497	Standard
	Sb	123	350.1	17.1	**0.0624**	0.012	19.0	ug/L	78	Standard
	Ba	135	341983.5	0.6	**148.3035**	0.748	0.5	ug/L	24	Standard
	Ce	140	2710163.1	0.4				ug/L	25	Standard
>	**Tb**	159	1034801.7	0.1				ug/L	939701	Standard
	Ho	165	19319.0	1.7				ug/L	10	Standard
	Tl	203	872.7	2.5	**0.0967**	0.002	2.5	ug/L	4	Standard
	Tl	205	843.4	5.5	**0.1098**	0.006	5.6	ug/L	2	Standard
	Pb	206	283131.4	1.0	**49.2257**	0.472	1.0	ug/L	240	Standard
	Pb	207	230315.5	1.3	**45.0642**	0.519	1.2	ug/L	198	Standard
	Pb	208	999318.0	0.9	**49.4947**	0.507	1.0	ug/L	883	Standard
	U	238	4899.8	1.7	**0.6576**	0.010	1.6	ug/L	3	Standard
>	**Bi**	209	563366.5	0.2				ug/L	523984	Standard

Sample ID: L1606120701
Report Date/Time: Wednesday, June 29, 2016 18:46:59
Page 1

Approved: July 04, 2016

Na	23	5.0	0.0	**48.0405**	0.159	0.3	mg/L	0	Standard
Mg	24	875.0	7.7	**0.8766**	0.075	8.6	mg/L	37	Standard
K	39	536.7	5.1	**0.8747**	0.054	6.2	mg/L	38	Standard
Ca	43	571.7	16.7	**27.5670**	4.925	17.9	mg/L	10	Standard
Fe	54	14788.8	1.9	**10.3138**	0.228	2.2	mg/L	179	Standard
Fe	57	4073.9	8.9	**10.4697**	1.059	10.1	mg/L	335	Standard
> **Sc-1**	45	73214.8	0.3				mg/L	61029	Standard
Cl	35	28264.3	0.5				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	3023.6	7.8				ug/L	1547	Standard
P	31	20599.0	2.2				ug/L	15286	Standard
S	34	2595.2	1.7				ug/L	1572	Standard
Sr	88	203.3	16.4				ug/L	73	Standard
C	12	130.0	40.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	150.0	43.7				mg/L	3	Standard
Dy	164	31532.4	1.2				mg/L	6	Standard
Ho-1	165	19319.0	1.7				mg/L	10	Standard
Er	166	17720.4	2.1				mg/L	13	Standard
I	127	25842.0	0.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.483	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.863	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120701
Report Date/Time: Wednesday, June 29, 2016 18:46:59
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.657
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.516
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120701
Report Date/Time: Wednesday, June 29, 2016 18:46:59
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120702
Sample Date/Time: Wednesday, June 29, 2016 18:47:54
Number of Replicates: 3
Autosampler Position: 248
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85381.8	3.5				ug/L	74699	Standard
	Be	9	623.3	3.3	**0.6272**	0.018	2.8	ug/L	8	Standard
	Al	27	4620203.8	1.9	**27.5041**	0.647	2.4	ug/L	1493	Standard
	Sc	45	75576.5	3.2				ug/L	61029	Standard
	Ti	47	6206.6	0.5	**14.3105**	0.192	1.3	ug/L	35	Standard
	V	51	134133.3	1.8	**15.4410**	0.153	1.0	ug/L	1435	Standard
	Cr	52	96500.3	0.7	**10.4416**	0.108	1.0	ug/L	7635	Standard
	Cr	53	11224.2	2.7	**10.3669**	0.448	4.3	ug/L	363	Standard
	Mn	55	3577332.4	1.7	**452.2740**	8.609	1.9	ug/L	747	Standard
	Co	59	47345.9	1.4	**6.2757**	0.086	1.4	ug/L	234	Standard
	Ni	60	24689.4	1.0	**11.4869**	0.101	0.9	ug/L	215	Standard
	Cu	65	29948.9	2.1	**14.0640**	0.244	1.7	ug/L	104	Standard
	Zn	66	164514.2	1.7	**132.1464**	1.033	0.8	ug/L	363	Standard
>	**Ge**	72	562390.3	1.6				ug/L	495142	Standard
	As	75	2776.0	3.7	**2.2412**	0.056	2.5	ug/L	-50	Standard
	Se	82	106.3	15.9	**0.7139**	0.124	17.3	ug/L	14	Standard
	Se-1	77	132.3	11.0	**0.9743**	0.190	19.5	ug/L	42	Standard
>	**Ga**	71	12940.5	1.1				mg/L	42	Standard
	Rb	85	250383.8	1.2				ug/L	17	Standard
	Y	89	695657.2	0.8				ug/L	417098	Standard
>	**Rh**	103	35.0	62.3				ug/L	13	Standard
	Mo	98	1645.8	3.3	**0.5188**	0.023	4.4	ug/L	11	Standard
	Ag	107	671.7	7.7	**0.0746**	0.007	9.4	ug/L	68	Standard
	Cd	111	2017.5	2.8	**0.8116**	0.028	3.5	mg/L	2	Standard
	Cd	114	4822.6	1.0	**0.7964**	0.005	0.7	ug/L	12	Standard
>	**In**	115	691832.3	1.1				ug/L	587441	Standard
	Sn	118	616.7	6.1	**-0.0050**	0.006	126.3	ug/L	497	Standard
	Sb	123	493.2	8.1	**0.0855**	0.008	9.1	ug/L	78	Standard
	Ba	135	281871.8	0.7	**116.9321**	1.732	1.5	ug/L	24	Standard
	Ce	140	1017603.0	1.2				ug/L	25	Standard
>	**Tb**	159	1062512.9	1.9				ug/L	939701	Standard
	Ho	165	14416.9	1.8				ug/L	10	Standard
	Tl	203	1059.0	3.0	**0.1166**	0.005	4.1	ug/L	4	Standard
	Tl	205	896.7	14.9	**0.1158**	0.019	16.8	ug/L	2	Standard
	Pb	206	100858.6	1.2	**17.3191**	0.532	3.1	ug/L	240	Standard
	Pb	207	81475.8	1.0	**15.7413**	0.458	2.9	ug/L	198	Standard
	Pb	208	356757.6	0.9	**17.4487**	0.446	2.6	ug/L	883	Standard
	U	238	6683.8	3.5	**0.8902**	0.048	5.4	ug/L	3	Standard
>	**Bi**	209	569106.4	1.9				ug/L	523984	Standard

Sample ID: L1606120702
Report Date/Time: Wednesday, June 29, 2016 18:50:11
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**31.5301**	27.313	86.6	mg/L	0	Standard
Mg	24	745.0	7.0	**0.7121**	0.061	8.5	mg/L	37	Standard
K	39	773.4	6.0	**1.2683**	0.120	9.5	mg/L	38	Standard
Ca	43	171.7	3.4	**6.2441**	0.546	8.7	mg/L	10	Standard
Fe	54	10724.1	1.5	**7.2104**	0.318	4.4	mg/L	179	Standard
Fe	57	2831.9	1.0	**6.7023**	0.187	2.8	mg/L	335	Standard
> **Sc-1**	45	75576.5	3.2				mg/L	61029	Standard
Cl	35	27596.4	1.5				ug/L	20195	Standard
Kr	83	4.0	66.1				ug/L	4	Standard
Br	81	8432.3	2.7				ug/L	1547	Standard
P	31	20674.1	3.5				ug/L	15286	Standard
S	34	2360.2	7.8				ug/L	1572	Standard
Sr	88	120.0	18.2				ug/L	73	Standard
C	12	350.0	15.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	106.7	43.3				mg/L	3	Standard
Dy	164	23142.4	1.5				mg/L	6	Standard
Ho-1	165	14416.9	1.8				mg/L	10	Standard
Er	166	13422.6	4.3				mg/L	13	Standard
I	127	174464.7	3.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		114.301	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.582	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120702
Report Date/Time: Wednesday, June 29, 2016 18:50:11
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	117.770
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.611
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120702
Report Date/Time: Wednesday, June 29, 2016 18:50:11
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120801
Sample Date/Time: Wednesday, June 29, 2016 18:51:05
Number of Replicates: 3
Autosampler Position: 249
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83030.4	3.3				ug/L	74699	Standard
	Be	9	1115.0	11.0	**1.1453**	0.091	8.0	ug/L	8	Standard
	Al	27	17314833.4	0.9	**106.0421**	2.666	2.5	ug/L	1493	Standard
	Sc	45	77183.1	5.1				ug/L	61029	Standard
	Ti	47	40548.2	1.5	**95.7436**	2.793	2.9	ug/L	35	Standard
	V	51	251782.0	1.4	**29.6946**	1.634	5.5	ug/L	1435	Standard
	Cr	52	243737.7	2.0	**28.4388**	0.710	2.5	ug/L	7635	Standard
	Cr	53	29872.8	2.3	**28.6239**	0.690	2.4	ug/L	363	Standard
	Mn	55	3685844.9	1.5	**474.5740**	27.736	5.8	ug/L	747	Standard
	Co	59	66271.6	1.6	**8.9620**	0.528	5.9	ug/L	234	Standard
	Ni	60	46026.2	2.6	**21.8948**	0.433	2.0	ug/L	215	Standard
	Cu	65	42002.8	1.5	**20.1164**	0.608	3.0	ug/L	104	Standard
	Zn	66	111824.0	2.6	**91.2379**	1.710	1.9	ug/L	363	Standard
>	**Ge**	72	553103.7	4.4				ug/L	495142	Standard
	As	75	5557.5	3.1	**4.5033**	0.093	2.1	ug/L	-50	Standard
	Se	82	102.2	5.1	**0.6977**	0.072	10.3	ug/L	14	Standard
	Se-1	77	134.3	12.1	**1.0314**	0.270	26.2	ug/L	42	Standard
>	**Ga**	71	24266.1	1.0				mg/L	42	Standard
	Rb	85	181545.0	2.7				ug/L	17	Standard
	Y	89	698995.8	3.0				ug/L	417098	Standard
>	**Rh**	103	88.3	6.5				ug/L	13	Standard
	Mo	98	2397.1	2.7	**0.8492**	0.049	5.8	ug/L	11	Standard
	Ag	107	1012.7	3.0	**0.1348**	0.009	6.5	ug/L	68	Standard
	Cd	111	911.2	3.1	**0.4060**	0.012	2.9	mg/L	2	Standard
	Cd	114	2080.2	4.3	**0.3830**	0.011	2.9	ug/L	12	Standard
>	**In**	115	618664.5	3.1				ug/L	587441	Standard
	Sn	118	1240.1	3.1	**0.1071**	0.013	12.0	ug/L	497	Standard
	Sb	123	1640.7	2.7	**0.3299**	0.019	5.8	ug/L	78	Standard
	Ba	135	534847.9	1.0	**248.3132**	9.882	4.0	ug/L	24	Standard
	Ce	140	1227932.9	1.4				ug/L	25	Standard
>	**Tb**	159	991058.9	3.3				ug/L	939701	Standard
	Ho	165	19152.1	0.5				ug/L	10	Standard
	Tl	203	1084.7	2.8	**0.1294**	0.004	2.8	ug/L	4	Standard
	Tl	205	976.7	6.4	**0.1361**	0.004	2.8	ug/L	2	Standard
	Pb	206	122644.8	1.6	**22.8364**	1.340	5.9	ug/L	240	Standard
	Pb	207	98294.5	1.0	**20.5914**	1.111	5.4	ug/L	198	Standard
	Pb	208	423360.3	0.4	**22.4444**	1.000	4.5	ug/L	883	Standard
	U	238	7556.2	1.8	**1.0902**	0.045	4.1	ug/L	3	Standard
>	**Bi**	209	526041.8	4.3				ug/L	523984	Standard

Sample ID: L1606120801
Report Date/Time: Wednesday, June 29, 2016 18:53:22
Page 1

Approved: July 04, 2016

Na	23	10.0	50.0	**92.8331**	50.540	54.4	mg/L	0	Standard
Mg	24	373.3	13.0	**0.3164**	0.050	15.9	mg/L	37	Standard
K	39	846.7	3.0	**1.3673**	0.064	4.7	mg/L	38	Standard
Ca	43	1881.8	8.7	**91.3781**	4.657	5.1	mg/L	10	Standard
Fe	54	16605.4	3.8	**11.0236**	0.942	8.5	mg/L	179	Standard
Fe	57	5247.6	6.9	**13.0290**	0.645	4.9	mg/L	335	Standard
Sc-1	45	77183.1	5.1				mg/L	61029	Standard
Cl	35	28165.5	0.9				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	3540.4	14.7				ug/L	1547	Standard
P	31	23239.5	5.6				ug/L	15286	Standard
S	34	2426.9	3.5				ug/L	1572	Standard
Sr	88	415.0	11.5				ug/L	73	Standard
C	12	173.3	6.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	66.7	60.6				mg/L	3	Standard
Dy	164	30988.1	3.0				mg/L	6	Standard
Ho-1	165	19152.1	0.5				mg/L	10	Standard
Er	166	17717.0	2.5				mg/L	13	Standard
I	127	27493.3	1.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		111.154	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		111.706	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120801
Report Date/Time: Wednesday, June 29, 2016 18:53:22
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	105.315
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	100.393
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120801
Report Date/Time: Wednesday, June 29, 2016 18:53:22
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120802
Sample Date/Time: Wednesday, June 29, 2016 18:54:17
Number of Replicates: 3
Autosampler Position: 250
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88923.6	2.8				ug/L	74699	Standard
	Be	9	661.7	11.4	**0.6387**	0.067	10.5	ug/L	8	Standard
	Al	27	15492795.6	1.0	**88.5762**	1.812	2.0	ug/L	1493	Standard
	Sc	45	80386.0	1.0				ug/L	61029	Standard
	Ti	47	20524.2	1.6	**45.7756**	1.470	3.2	ug/L	35	Standard
	V	51	220306.5	1.3	**24.5176**	0.534	2.2	ug/L	1435	Standard
	Cr	52	643370.5	1.6	**72.5457**	2.733	3.8	ug/L	7635	Standard
	Cr	53	78035.6	3.0	**71.2237**	3.200	4.5	ug/L	363	Standard
	Mn	55	3332205.3	1.0	**405.3306**	7.931	2.0	ug/L	747	Standard
	Co	59	49008.5	0.3	**6.2510**	0.159	2.5	ug/L	234	Standard
	Ni	60	36936.6	2.2	**16.5952**	0.689	4.2	ug/L	215	Standard
	Cu	65	48506.2	1.8	**21.9786**	0.680	3.1	ug/L	104	Standard
	Zn	66	431997.7	0.4	**334.6356**	8.273	2.5	ug/L	363	Standard
>	**Ge**	72	584593.6	2.3				ug/L	495142	Standard
	As	75	4706.8	0.7	**3.6196**	0.063	1.8	ug/L	-50	Standard
	Se	82	170.0	3.4	**1.1737**	0.017	1.5	ug/L	14	Standard
	Se-1	77	180.7	11.7	**1.4743**	0.259	17.6	ug/L	42	Standard
>	**Ga**	71	19881.4	1.3				mg/L	42	Standard
	Rb	85	200103.0	1.8				ug/L	17	Standard
	Y	89	657665.4	0.6				ug/L	417098	Standard
>	**Rh**	103	171.7	7.3				ug/L	13	Standard
	Mo	98	1216.9	1.5	**0.3934**	0.011	2.8	ug/L	11	Standard
	Ag	107	1006.0	3.7	**0.1221**	0.007	5.4	ug/L	68	Standard
	Cd	111	4519.9	2.1	**1.8818**	0.049	2.6	mg/L	2	Standard
	Cd	114	11256.2	1.6	**1.9163**	0.030	1.6	ug/L	12	Standard
>	**In**	115	672127.4	1.5				ug/L	587441	Standard
	Sn	118	1105.0	15.3	**0.0707**	0.026	36.7	ug/L	497	Standard
	Sb	123	229.1	10.3	**0.0387**	0.005	12.5	ug/L	78	Standard
	Ba	135	361495.2	1.1	**154.3886**	3.928	2.5	ug/L	24	Standard
	Ce	140	936346.8	0.9				ug/L	25	Standard
>	**Tb**	159	1070983.4	3.2				ug/L	939701	Standard
	Ho	165	13711.2	2.1				ug/L	10	Standard
	Tl	203	1023.7	3.8	**0.1103**	0.006	5.8	ug/L	4	Standard
	Tl	205	995.0	4.8	**0.1256**	0.007	5.2	ug/L	2	Standard
	Pb	206	1451242.4	0.5	**244.9779**	5.608	2.3	ug/L	240	Standard
	Pb	207	1189055.7	0.7	**225.9168**	5.756	2.5	ug/L	198	Standard
	Pb	208	5216722.9	1.2	**250.8847**	7.037	2.8	ug/L	883	Standard
	U	238	9013.7	1.2	**1.1768**	0.034	2.9	ug/L	3	Standard
>	**Bi**	209	581076.6	1.9				ug/L	523984	Standard

Sample ID: L1606120802
Report Date/Time: Wednesday, June 29, 2016 18:56:34
Page 1

Approved: July 04, 2016

Na	23	6.7	173.2	**57.8631**	100.213	173.2	mg/L	0	Standard
Mg	24	553.3	14.7	**0.4776**	0.083	17.4	mg/L	37	Standard
K	39	1005.0	9.9	**1.5724**	0.167	10.6	mg/L	38	Standard
Ca	43	1240.1	6.3	**56.9237**	3.160	5.6	mg/L	10	Standard
Fe	54	17498.7	0.5	**11.1262**	0.159	1.4	mg/L	179	Standard
Fe	57	5124.2	2.5	**12.1481**	0.285	2.3	mg/L	335	Standard
Sc-1	45	80386.0	1.0				mg/L	61029	Standard
Cl	35	28790.7	0.3				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	8875.9	6.3				ug/L	1547	Standard
P	31	20292.0	8.4				ug/L	15286	Standard
S	34	2478.5	4.1				ug/L	1572	Standard
Sr	88	438.3	2.4				ug/L	73	Standard
C	12	323.3	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	136.7	34.6				mg/L	3	Standard
Dy	164	22618.6	1.5				mg/L	6	Standard
Ho-1	165	13711.2	2.1				mg/L	10	Standard
Er	166	12578.6	1.5				mg/L	13	Standard
I	127	73652.8	8.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		119.043	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		118.066	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120802
Report Date/Time: Wednesday, June 29, 2016 18:56:34
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	114.416
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	110.896
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120802
Report Date/Time: Wednesday, June 29, 2016 18:56:34
Page 3

Approved: July 04, 2016

Pb 206 Upper, S, EEE	Pb	206
Pb 207 Upper, S, EEE	Pb	207
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129101
Sample Date/Time: Wednesday, June 29, 2016 18:57:28
Number of Replicates: 3
Autosampler Position: 251
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89580.8	1.7				ug/L	74699	Standard
	Be	9	360.0	14.7	**0.3489**	0.053	15.3	ug/L	8	Standard
	Al	27	3187676.8	1.0	**18.0772**	0.341	1.9	ug/L	1493	Standard
	Sc	45	73445.9	1.6				ug/L	61029	Standard
	Ti	47	23807.7	6.7	**54.3513**	2.712	5.0	ug/L	35	Standard
	V	51	62405.7	2.9	**6.9715**	0.072	1.0	ug/L	1435	Standard
	Cr	52	100378.5	1.9	**10.7261**	0.052	0.5	ug/L	7635	Standard
	Cr	53	11571.1	3.7	**10.5314**	0.357	3.4	ug/L	363	Standard
	Mn	55	3995310.3	0.7	**497.7426**	6.886	1.4	ug/L	747	Standard
	Co	59	12689.3	1.5	**1.6298**	0.018	1.1	ug/L	234	Standard
	Ni	60	15702.5	1.8	**7.1516**	0.026	0.4	ug/L	215	Standard
	Cu	65	1030601.6	2.5	**480.2115**	2.610	0.5	ug/L	104	Standard
	Zn	66	647092.9	2.6	**513.3926**	3.363	0.7	ug/L	363	Standard
>	**Ge**	72	570780.0	2.0				ug/L	495142	Standard
	As	75	23305.2	1.3	**18.1104**	0.138	0.8	ug/L	-50	Standard
	Se	82	481.4	2.5	**3.6678**	0.121	3.3	ug/L	14	Standard
	Se-1	77	377.3	9.1	**3.8600**	0.402	10.4	ug/L	42	Standard
>	**Ga**	71	43582.7	4.5				mg/L	42	Standard
	Rb	85	26201.0	3.0				ug/L	17	Standard
	Y	89	551000.5	1.0				ug/L	417098	Standard
>	**Rh**	103	178.3	18.7				ug/L	13	Standard
	Mo	98	21134.4	1.3	**6.6476**	0.191	2.9	ug/L	11	Standard
	Ag	107	15204.6	2.3	**1.9441**	0.032	1.6	ug/L	68	Standard
	Cd	111	1466.1	5.3	**0.5801**	0.025	4.4	mg/L	2	Standard
	Cd	114	3712.7	3.5	**0.6050**	0.007	1.1	ug/L	12	Standard
>	**In**	115	700391.9	2.5				ug/L	587441	Standard
	Sn	118	18187.6	2.5	**2.5189**	0.131	5.2	ug/L	497	Standard
	Sb	123	7709.0	2.8	**1.3807**	0.022	1.6	ug/L	78	Standard
	Ba	135	971846.2	1.7	**398.2918**	4.112	1.0	ug/L	24	Standard
	Ce	140	138638.2	1.7				ug/L	25	Standard
>	**Tb**	159	1070571.9	3.6				ug/L	939701	Standard
	Ho	165	2720.2	8.8				ug/L	10	Standard
	Tl	203	657.3	30.6	**0.0422**	0.014	32.0	ug/L	4	Standard
	Tl	205	563.3	26.0	**0.0436**	0.011	26.1	ug/L	2	Standard
	Pb	206	89328.6	2.2	**9.1434**	0.147	1.6	ug/L	240	Standard
	Pb	207	72922.4	1.4	**8.3969**	0.107	1.3	ug/L	198	Standard
	Pb	208	316979.9	0.6	**9.2432**	0.217	2.4	ug/L	883	Standard
	U	238	78871.8	1.7	**6.3120**	0.222	3.5	ug/L	3	Standard
>	**Bi**	209	950965.0	2.0				ug/L	523984	Standard

Sample ID: L1606129101
Report Date/Time: Wednesday, June 29, 2016 18:59:45
Page 1

Approved: July 04, 2016

Na	23	5.0	173.2	**48.7719**	84.467	173.2	mg/L	0	Standard
Mg	24	613.3	9.8	**0.5934**	0.073	12.3	mg/L	37	Standard
K	39	323.3	7.3	**0.4798**	0.044	9.2	mg/L	38	Standard
Ca	43	198.3	14.8	**7.8949**	1.614	20.4	mg/L	10	Standard
Fe	54	8830.8	5.5	**6.0856**	0.392	6.4	mg/L	179	Standard
Fe	57	2530.2	2.9	**6.0730**	0.228	3.8	mg/L	335	Standard
Sc-1	45	73445.9	1.6				mg/L	61029	Standard
Cl	35	25952.9	0.3				ug/L	20195	Standard
Kr	83	2.0	100.0				ug/L	4	Standard
Br	81	8005.5	5.5				ug/L	1547	Standard
P	31	19060.3	1.8				ug/L	15286	Standard
S	34	1756.8	1.0				ug/L	1572	Standard
Sr	88	518.3	2.8				ug/L	73	Standard
C	12	1226.7	9.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	1270.1	11.0				mg/L	3	Standard
Dy	164	4157.3	8.4				mg/L	6	Standard
Ho-1	165	2720.2	8.8				mg/L	10	Standard
Er	166	2643.6	8.4				mg/L	13	Standard
I	127	191520.8	4.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		119.923	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.276	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606129101
Report Date/Time: Wednesday, June 29, 2016 18:59:45
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	119.228
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	181.487
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Cu 65 Upper, S, EEE	Cu	65	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606129101
Report Date/Time: Wednesday, June 29, 2016 18:59:45
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135	
Bi 209 Int Std for sample	Bi	209	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138101
Sample Date/Time: Wednesday, June 29, 2016 19:00:40
Number of Replicates: 3
Autosampler Position: 252
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96098.8	1.2				ug/L	74699	Standard
	Be	9	1650.1	8.9	**1.4642**	0.124	8.5	ug/L	8	Standard
	Al	27	23502208.7	0.7	**124.3024**	1.056	0.8	ug/L	1493	Standard
	Sc	45	81510.4	1.4				ug/L	61029	Standard
	Ti	47	9357.2	0.5	**21.1737**	0.376	1.8	ug/L	35	Standard
	V	51	257600.4	1.3	**29.2098**	0.781	2.7	ug/L	1435	Standard
	Cr	52	224787.2	0.6	**25.1231**	0.400	1.6	ug/L	7635	Standard
	Cr	53	27600.1	3.0	**25.4059**	0.410	1.6	ug/L	363	Standard
	Mn	55	5543699.6	0.4	**686.2289**	8.924	1.3	ug/L	747	Standard
	Co	59	117640.6	0.9	**15.3216**	0.399	2.6	ug/L	234	Standard
	Ni	60	64249.3	0.8	**29.4562**	0.340	1.2	ug/L	215	Standard
	Cu	65	57365.7	1.4	**26.4629**	0.232	0.9	ug/L	104	Standard
	Zn	66	103406.0	1.2	**81.1457**	1.055	1.3	ug/L	363	Standard
>	**Ge**	72	574477.7	1.7				ug/L	495142	Standard
	As	75	12246.4	0.5	**9.4841**	0.110	1.2	ug/L	-50	Standard
	Se	82	125.1	3.5	**0.8446**	0.024	2.8	ug/L	14	Standard
	Se-1	77	164.3	2.5	**1.3176**	0.070	5.3	ug/L	42	Standard
>	**Ga**	71	29587.2	2.2				mg/L	42	Standard
	Rb	85	313937.1	2.6				ug/L	17	Standard
	Y	89	822636.0	0.8				ug/L	417098	Standard
>	**Rh**	103	38.3	49.4				ug/L	13	Standard
	Mo	98	757.2	4.8	**0.2454**	0.010	4.2	ug/L	11	Standard
	Ag	107	953.4	1.4	**0.1164**	0.003	2.3	ug/L	68	Standard
	Cd	111	1588.5	0.6	**0.6638**	0.009	1.4	mg/L	2	Standard
	Cd	114	3783.2	0.8	**0.6501**	0.009	1.4	ug/L	12	Standard
>	**In**	115	664448.0	1.0				ug/L	587441	Standard
	Sn	118	643.3	5.2	**0.0026**	0.004	152.2	ug/L	497	Standard
	Sb	123	269.8	13.5	**0.0469**	0.007	15.6	ug/L	78	Standard
	Ba	135	662128.9	0.5	**286.0118**	2.987	1.0	ug/L	24	Standard
	Ce	140	1687892.1	1.1				ug/L	25	Standard
>	**Tb**	159	1062435.8	0.8				ug/L	939701	Standard
	Ho	165	27182.7	2.2				ug/L	10	Standard
	Tl	203	1364.1	4.3	**0.1521**	0.007	4.3	ug/L	4	Standard
	Tl	205	1335.1	6.0	**0.1737**	0.010	5.7	ug/L	2	Standard
	Pb	206	284182.0	0.3	**49.3998**	0.214	0.4	ug/L	240	Standard
	Pb	207	234533.7	1.0	**45.8828**	0.448	1.0	ug/L	198	Standard
	Pb	208	1023299.4	0.1	**50.6751**	0.295	0.6	ug/L	883	Standard
	U	238	6312.3	2.1	**0.8483**	0.014	1.7	ug/L	3	Standard
>	**Bi**	209	563477.4	0.7				ug/L	523984	Standard

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:02:57
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**28.8214**	24.963	86.6	mg/L	0	Standard
Mg	24	151.7	18.2	**0.0816**	0.027	33.5	mg/L	37	Standard
K	39	1206.7	9.1	**1.8839**	0.198	10.5	mg/L	38	Standard
Ca	43	230.0	38.1	**8.3188**	4.002	48.1	mg/L	10	Standard
Fe	54	27991.0	1.6	**17.6350**	0.523	3.0	mg/L	179	Standard
Fe	57	7366.8	1.7	**17.6771**	0.511	2.9	mg/L	335	Standard
Sc-1	45	81510.4	1.4				mg/L	61029	Standard
Cl	35	28359.2	2.3				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	4510.7	3.7				ug/L	1547	Standard
P	31	20567.3	1.1				ug/L	15286	Standard
S	34	2398.5	2.1				ug/L	1572	Standard
Sr	88	138.3	12.7				ug/L	73	Standard
C	12	220.0	4.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	113.3	35.7				mg/L	3	Standard
Dy	164	44981.5	2.0				mg/L	6	Standard
Ho-1	165	27182.7	2.2				mg/L	10	Standard
Er	166	24606.7	4.9				mg/L	13	Standard
I	127	24668.4	3.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		128.648	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		116.023	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:02:57
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.109
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.537
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:02:57
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138101
Sample Date/Time: Wednesday, June 29, 2016 19:03:51
Number of Replicates: 3
Autosampler Position: 253
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89738.3	1.4				ug/L	74699	Standard
	Be	9	41.7	6.9	**0.0470**	0.003	6.5	ug/L	8	Standard
	Al	27	505536.2	3.8	**2.8498**	0.087	3.1	ug/L	1493	Standard
	Sc	45	72476.3	1.0				ug/L	61029	Standard
	Ti	47	238.7	3.3	**0.4178**	0.015	3.5	ug/L	35	Standard
	V	51	6994.4	2.1	**0.5793**	0.021	3.7	ug/L	1435	Standard
	Cr	52	13067.6	0.8	**0.4515**	0.011	2.5	ug/L	7635	Standard
	Cr	53	1060.0	5.3	**0.6298**	0.052	8.3	ug/L	363	Standard
	Mn	55	117506.1	2.5	**13.9721**	0.301	2.2	ug/L	747	Standard
	Co	59	2633.2	1.9	**0.2964**	0.006	2.1	ug/L	234	Standard
	Ni	60	1579.1	3.6	**0.5819**	0.030	5.1	ug/L	215	Standard
	Cu	65	1415.4	1.9	**0.5327**	0.017	3.2	ug/L	104	Standard
	Zn	66	4415.0	0.5	**2.9272**	0.025	0.9	ug/L	363	Standard
>	**Ge**	72	591804.3	0.7				ug/L	495142	Standard
	As	75	188.6	12.3	**0.2004**	0.017	8.4	ug/L	-50	Standard
	Se	82	17.7	38.3	**-0.0030**	0.051	1669.1	ug/L	14	Standard
	Se-1	77	53.3	24.5	**-0.0120**	0.145	1206.4	ug/L	42	Standard
>	**Ga**	71	651.7	3.8				mg/L	42	Standard
	Rb	85	6646.5	2.9				ug/L	17	Standard
	Y	89	496582.4	1.4				ug/L	417098	Standard
>	**Rh**	103	10.0	86.6				ug/L	13	Standard
	Mo	98	63.7	13.0	**0.0149**	0.003	17.4	ug/L	11	Standard
	Ag	107	138.0	7.0	**0.0054**	0.001	21.7	ug/L	68	Standard
	Cd	111	41.6	25.2	**0.0093**	0.004	47.8	mg/L	2	Standard
	Cd	114	78.4	6.5	**0.0111**	0.001	8.0	ug/L	12	Standard
>	**In**	115	680653.4	0.9				ug/L	587441	Standard
	Sn	118	558.3	7.0	**-0.0122**	0.005	43.2	ug/L	497	Standard
	Sb	123	143.1	16.8	**0.0222**	0.004	19.8	ug/L	78	Standard
	Ba	135	14129.9	0.6	**5.9423**	0.035	0.6	ug/L	24	Standard
	Ce	140	36747.8	1.4				ug/L	25	Standard
>	**Tb**	159	1046560.0	0.5				ug/L	939701	Standard
	Ho	165	620.0	6.6				ug/L	10	Standard
	Tl	203	40.7	9.9	**0.0027**	0.000	16.9	ug/L	4	Standard
	Tl	205	38.3	19.9	**0.0050**	0.001	18.8	ug/L	2	Standard
	Pb	206	6390.7	0.3	**1.0069**	0.002	0.2	ug/L	240	Standard
	Pb	207	5305.6	1.7	**0.9374**	0.017	1.8	ug/L	198	Standard
	Pb	208	22358.8	1.4	**1.0005**	0.020	2.0	ug/L	883	Standard
	U	238	147.3	10.0	**0.0149**	0.002	12.5	ug/L	3	Standard
>	**Bi**	209	581458.5	0.4				ug/L	523984	Standard

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:06:08
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	36.7	41.7	**-0.0251**	0.017	67.5	mg/L	37	Standard
K	39	73.3	27.6	**0.0225**	0.038	170.4	mg/L	38	Standard
Ca	43	25.0	40.0	**-1.1927**	0.538	45.1	mg/L	10	Standard
Fe	54	777.1	4.9	**0.4166**	0.030	7.2	mg/L	179	Standard
Fe	57	423.3	1.8	**0.1373**	0.012	9.1	mg/L	335	Standard
> **Sc-1**	45	72476.3	1.0				mg/L	61029	Standard
Cl	35	26907.2	2.4				ug/L	20195	Standard
Kr	83	1.7	124.9				ug/L	4	Standard
Br	81	2140.2	2.1				ug/L	1547	Standard
P	31	19854.7	4.5				ug/L	15286	Standard
S	34	2031.8	8.8				ug/L	1572	Standard
Sr	88	61.7	24.8				ug/L	73	Standard
C	12	66.7	82.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	1031.0	18.9				mg/L	6	Standard
Ho-1	165	620.0	6.6				mg/L	10	Standard
Er	166	536.7	12.4				mg/L	13	Standard
I	127	3175.3	27.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		120.134	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		119.522	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:06:08
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		115.868
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.969
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138102
Sample Date/Time: Wednesday, June 29, 2016 19:07:03
Number of Replicates: 3
Autosampler Position: 254
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93009.7	1.6				ug/L	74699	Standard
	Be	9	1248.4	3.0	**1.1470**	0.051	4.5	ug/L	8	Standard
	Al	27	40124336.3	0.4	**219.2936**	2.793	1.3	ug/L	1493	Standard
	Sc	45	83437.5	1.7				ug/L	61029	Standard
	Ti	47	9733.8	0.5	**22.4436**	0.290	1.3	ug/L	35	Standard
	V	51	219254.4	2.3	**25.2954**	0.490	1.9	ug/L	1435	Standard
	Cr	52	167385.2	0.7	**18.8093**	0.105	0.6	ug/L	7635	Standard
	Cr	53	20240.2	3.5	**18.8976**	0.529	2.8	ug/L	363	Standard
	Mn	55	6543259.4	1.4	**825.1579**	15.855	1.9	ug/L	747	Standard
	Co	59	86492.3	2.3	**11.4640**	0.252	2.2	ug/L	234	Standard
	Ni	60	58233.4	1.3	**27.1864**	0.121	0.4	ug/L	215	Standard
	Cu	65	35144.1	1.8	**16.4764**	0.210	1.3	ug/L	104	Standard
	Zn	66	61921.6	1.7	**49.3206**	0.509	1.0	ug/L	363	Standard
>	**Ge**	72	563884.1	0.9				ug/L	495142	Standard
	As	75	5876.5	0.9	**4.6669**	0.080	1.7	ug/L	-50	Standard
	Se	82	62.7	20.2	**0.3638**	0.101	27.7	ug/L	14	Standard
	Se-1	77	116.3	8.0	**0.7757**	0.099	12.7	ug/L	42	Standard
>	**Ga**	71	24224.4	1.6				mg/L	42	Standard
	Rb	85	231886.4	1.9				ug/L	17	Standard
	Y	89	788603.8	0.2				ug/L	417098	Standard
>	**Rh**	103	76.7	50.7				ug/L	13	Standard
	Mo	98	590.2	6.7	**0.1909**	0.011	6.0	ug/L	11	Standard
	Ag	107	590.3	14.1	**0.0675**	0.010	15.2	ug/L	68	Standard
	Cd	111	402.0	10.7	**0.1627**	0.016	9.9	mg/L	2	Standard
	Cd	114	896.6	5.0	**0.1531**	0.007	4.5	ug/L	12	Standard
>	**In**	115	661468.7	1.5				ug/L	587441	Standard
	Sn	118	563.3	17.0	**-0.0090**	0.016	173.2	ug/L	497	Standard
	Sb	123	168.4	11.6	**0.0279**	0.004	15.0	ug/L	78	Standard
	Ba	135	766980.0	0.9	**332.7995**	3.011	0.9	ug/L	24	Standard
	Ce	140	1513996.2	0.9				ug/L	25	Standard
>	**Tb**	159	1057537.7	1.8				ug/L	939701	Standard
	Ho	165	25513.1	1.6				ug/L	10	Standard
	Tl	203	1321.7	3.8	**0.1489**	0.005	3.2	ug/L	4	Standard
	Tl	205	1201.7	6.5	**0.1579**	0.009	5.7	ug/L	2	Standard
	Pb	206	69846.7	2.1	**12.2126**	0.140	1.1	ug/L	240	Standard
	Pb	207	56032.1	2.8	**11.0205**	0.188	1.7	ug/L	198	Standard
	Pb	208	243026.5	0.6	**12.1032**	0.059	0.5	ug/L	883	Standard
	U	238	7981.8	0.7	**1.0851**	0.004	0.3	ug/L	3	Standard
>	**Bi**	209	557709.9	1.1				ug/L	523984	Standard

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:09:20
Page 1

Approved: July 04, 2016

Na	23	6.7	86.6	**56.4197**	48.874	86.6	mg/L	0	Standard
Mg	24	238.3	32.7	**0.1607**	0.077	48.1	mg/L	37	Standard
K	39	688.3	8.8	**0.9980**	0.083	8.3	mg/L	38	Standard
Ca	43	911.7	12.3	**39.6249**	5.390	13.6	mg/L	10	Standard
Fe	54	20800.6	0.9	**12.7625**	0.173	1.4	mg/L	179	Standard
Fe	57	6012.9	2.1	**13.8749**	0.248	1.8	mg/L	335	Standard
Sc-1	45	83437.5	1.7				mg/L	61029	Standard
Cl	35	29332.4	1.9				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2603.6	11.5				ug/L	1547	Standard
P	31	21632.1	2.1				ug/L	15286	Standard
S	34	2746.9	3.7				ug/L	1572	Standard
Sr	88	253.3	10.9				ug/L	73	Standard
C	12	116.7	27.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	33.3	45.8				mg/L	3	Standard
Dy	164	42471.8	1.0				mg/L	6	Standard
Ho-1	165	25513.1	1.6				mg/L	10	Standard
Er	166	22212.9	0.4				mg/L	13	Standard
I	127	23259.5	0.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		124.513	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		113.883	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:09:20
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.602
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.436
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:09:20
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138102
Sample Date/Time: Wednesday, June 29, 2016 19:10:14
Number of Replicates: 3
Autosampler Position: 255
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85837.5	1.0				ug/L	74699	Standard
	Be	9	38.3	19.9	**0.0455**	0.007	16.1	ug/L	8	Standard
	Al	27	829500.0	3.1	**4.8979**	0.105	2.1	ug/L	1493	Standard
	Sc	45	72806.2	1.9				ug/L	61029	Standard
	Ti	47	228.0	6.2	**0.3999**	0.033	8.1	ug/L	35	Standard
	V	51	5937.2	2.5	**0.4694**	0.017	3.6	ug/L	1435	Standard
	Cr	52	11356.6	1.4	**0.2725**	0.020	7.2	ug/L	7635	Standard
	Cr	53	790.0	13.9	**0.3928**	0.096	24.5	ug/L	363	Standard
	Mn	55	140399.5	1.1	**16.9108**	0.252	1.5	ug/L	747	Standard
	Co	59	1986.1	4.2	**0.2171**	0.011	4.8	ug/L	234	Standard
	Ni	60	1398.7	1.4	**0.5081**	0.009	1.8	ug/L	215	Standard
	Cu	65	887.4	6.1	**0.2997**	0.027	8.9	ug/L	104	Standard
	Zn	66	3530.4	0.8	**2.2796**	0.035	1.5	ug/L	363	Standard
>	**Ge**	72	585343.3	0.6				ug/L	495142	Standard
	As	75	60.2	40.7	**0.1052**	0.019	17.9	ug/L	-50	Standard
	Se	82	19.5	23.7	**0.0127**	0.037	290.2	ug/L	14	Standard
	Se-1	77	54.0	19.3	**0.0032**	0.122	3775.1	ug/L	42	Standard
>	**Ga**	71	611.7	10.2				mg/L	42	Standard
	Rb	85	4939.1	4.6				ug/L	17	Standard
	Y	89	509652.0	1.0				ug/L	417098	Standard
>	**Rh**	103	5.0	100.0				ug/L	13	Standard
	Mo	98	44.1	7.7	**0.0084**	0.001	14.8	ug/L	11	Standard
	Ag	107	136.0	8.3	**0.0049**	0.002	33.6	ug/L	68	Standard
	Cd	111	13.6	15.2	**-0.0024**	0.001	32.7	mg/L	2	Standard
	Cd	114	34.6	39.9	**0.0036**	0.002	62.3	ug/L	12	Standard
>	**In**	115	689569.5	1.2				ug/L	587441	Standard
	Sn	118	635.0	8.5	**-0.0021**	0.008	369.1	ug/L	497	Standard
	Sb	123	120.9	7.4	**0.0179**	0.002	10.7	ug/L	78	Standard
	Ba	135	16312.1	1.7	**6.7731**	0.043	0.6	ug/L	24	Standard
	Ce	140	31461.0	1.2				ug/L	25	Standard
>	**Tb**	159	1048084.0	0.7				ug/L	939701	Standard
	Ho	165	503.3	20.9				ug/L	10	Standard
	Tl	203	28.0	9.4	**0.0013**	0.000	21.0	ug/L	4	Standard
	Tl	205	13.3	21.7	**0.0019**	0.000	19.6	ug/L	2	Standard
	Pb	206	1789.4	1.9	**0.2333**	0.006	2.6	ug/L	240	Standard
	Pb	207	1416.1	2.7	**0.2012**	0.009	4.3	ug/L	198	Standard
	Pb	208	6428.2	1.8	**0.2375**	0.004	1.7	ug/L	883	Standard
	U	238	157.3	5.4	**0.0164**	0.001	7.5	ug/L	3	Standard
>	**Bi**	209	576421.5	0.5				ug/L	523984	Standard

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:12:31
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	65.0	13.3	**0.0052**	0.008	161.5	mg/L	37	Standard
K	39	56.7	10.2	**-0.0092**	0.009	98.8	mg/L	38	Standard
Ca	43	46.7	16.4	**-0.0469**	0.452	963.7	mg/L	10	Standard
Fe	54	615.5	5.2	**0.2993**	0.025	8.2	mg/L	179	Standard
Fe	57	410.0	10.0	**0.0928**	0.096	103.4	mg/L	335	Standard
> **Sc-1**	45	72806.2	1.9				mg/L	61029	Standard
Cl	35	29355.1	3.4				ug/L	20195	Standard
Kr	83	2.3	24.7				ug/L	4	Standard
Br	81	1910.1	15.3				ug/L	1547	Standard
P	31	20493.9	2.3				ug/L	15286	Standard
S	34	2463.5	6.9				ug/L	1572	Standard
Sr	88	83.3	42.1				ug/L	73	Standard
C	12	50.0	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	991.0	1.5				mg/L	6	Standard
Ho-1	165	503.3	20.9				mg/L	10	Standard
Er	166	466.7	4.5				mg/L	13	Standard
I	127	2401.9	23.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		114.911	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		118.217	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:12:31
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.385
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.007
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 19:13:27
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83999.1	3.6				ug/L	74699	Standard
	Be	9	57157.7	3.6	**57.7443**	0.451	0.8	ug/L	8	Standard
	Al	27	8858742.4	2.0	**53.6133**	0.921	1.7	ug/L	1493	Standard
	Sc	45	70122.0	3.5				ug/L	61029	Standard
	Ti	47	47445.5	0.9	**109.3528**	3.353	3.1	ug/L	35	Standard
	V	51	471838.1	3.0	**54.4025**	0.466	0.9	ug/L	1435	Standard
	Cr	52	466548.5	1.7	**54.0130**	1.261	2.3	ug/L	7635	Standard
	Cr	53	57204.5	1.8	**53.7881**	1.637	3.0	ug/L	363	Standard
	Mn	55	422873.4	1.8	**52.9497**	1.131	2.1	ug/L	747	Standard
	Co	59	420575.5	1.8	**55.6399**	1.134	2.0	ug/L	234	Standard
	Ni	60	114954.2	1.8	**53.5507**	1.121	2.1	ug/L	215	Standard
	Cu	65	113826.7	2.0	**53.3592**	1.065	2.0	ug/L	104	Standard
	Zn	66	66150.5	1.8	**52.4882**	1.217	2.3	ug/L	363	Standard
>	**Ge**	72	566689.7	3.9				ug/L	495142	Standard
	As	75	66474.1	0.9	**51.9549**	1.668	3.2	ug/L	-50	Standard
	Se	82	6669.2	0.5	**53.0020**	1.808	3.4	ug/L	14	Standard
	Se-1	77	4569.4	1.3	**54.0883**	1.963	3.6	ug/L	42	Standard
>	**Ga**	71	60.0	25.0				mg/L	42	Standard
	Rb	85	973.4	4.9				ug/L	17	Standard
	Y	89	482055.4	0.8				ug/L	417098	Standard
>	**Rh**	103	21.7	35.3				ug/L	13	Standard
	Mo	98	311127.6	1.9	**101.8685**	1.475	1.4	ug/L	11	Standard
	Ag	107	381455.1	1.8	**51.0679**	0.625	1.2	ug/L	68	Standard
	Cd	111	122240.5	1.1	**51.0178**	0.254	0.5	mg/L	2	Standard
	Cd	114	297421.0	1.0	**50.6157**	0.803	1.6	ug/L	12	Standard
>	**In**	115	673075.7	0.6				ug/L	587441	Standard
	Sn	118	341770.5	1.5	**50.9666**	0.480	0.9	ug/L	497	Standard
	Sb	123	266115.5	1.9	**49.7407**	0.724	1.5	ug/L	78	Standard
	Ba	135	114468.8	0.7	**48.7952**	0.078	0.2	ug/L	24	Standard
	Ce	140	298.3	40.9				ug/L	25	Standard
>	**Tb**	159	1044481.4	1.5				ug/L	939701	Standard
	Ho	165	18.3	83.3				ug/L	10	Standard
	Tl	203	467319.7	1.5	**52.0663**	0.128	0.2	ug/L	4	Standard
	Tl	205	421316.6	1.5	**54.0865**	0.384	0.7	ug/L	2	Standard
	Pb	206	294059.2	2.1	**50.4836**	0.259	0.5	ug/L	240	Standard
	Pb	207	260305.8	2.4	**50.2976**	0.406	0.8	ug/L	198	Standard
	Pb	208	1086866.5	0.8	**53.1659**	0.550	1.0	ug/L	883	Standard
	U	238	395799.2	0.7	**52.8135**	0.760	1.4	ug/L	3	Standard
>	**Bi**	209	570529.8	1.7				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:15:44
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4777.4	3.9	**5.3091**	0.347	6.5	mg/L	37	Standard
K	39	2250.2	3.7	**4.2173**	0.198	4.7	mg/L	38	Standard
Ca	43	128.3	12.5	**4.5532**	1.080	23.7	mg/L	10	Standard
Fe	54	7377.7	1.8	**5.3086**	0.144	2.7	mg/L	179	Standard
Fe	57	2055.1	1.5	**5.0082**	0.127	2.5	mg/L	335	Standard
> **Sc-1**	45	70122.0	3.5				mg/L	61029	Standard
Cl	35	30001.0	0.9				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	2183.5	6.8				ug/L	1547	Standard
P	31	22995.8	1.5				ug/L	15286	Standard
S	34	2830.3	2.9				ug/L	1572	Standard
Sr	88	66.7	24.1				ug/L	73	Standard
C	12	80.0	50.0				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	9.0	108.0				mg/L	6	Standard
Ho-1	165	18.3	83.3				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	1968.5	24.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	115.489		
Al	27	107.227		
Sc	45			
Ti	47	109.353		
V	51	108.805		
Cr	52	108.026		
Cr	53			
Mn	55	105.899		
Co	59	111.280		
Ni	60	107.101		
Cu	65	106.718		
Zn	66	104.976		
> Ge	72		114.450	
As	75	103.910		
Se	82	106.004		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:15:44
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.868		
Ag	107	102.136		
Cd	111	102.036		
Cd	114			
In	115		114.578	
Sn	118	101.933		
Sb	123	99.481		
Ba	135	97.590		
Ce	140			
Tb	159			
Ho	165			
Tl	203	104.133		
Tl	205			
Pb	206			
Pb	207			
Pb	208	106.332		
U	238	105.627		
Bi	209		108.883	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:15:44
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 19:16:39
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83692.3	2.8				ug/L	74699	Standard
	Be	9	18.3	15.7	**0.0263**	0.003	12.9	ug/L	8	Standard
	Al	27	3213.7	25.4	**0.0061**	0.004	71.9	ug/L	1493	Standard
	Sc	45	70726.5	3.0				ug/L	61029	Standard
	Ti	47	41.0	12.2	**-0.0158**	0.012	73.1	ug/L	35	Standard
	V	51	1525.1	1.9	**-0.0214**	0.005	22.4	ug/L	1435	Standard
	Cr	52	7659.9	0.9	**-0.1324**	0.015	11.0	ug/L	7635	Standard
	Cr	53	410.0	16.0	**0.0526**	0.061	116.5	ug/L	363	Standard
	Mn	55	997.7	7.3	**-0.0244**	0.010	41.0	ug/L	747	Standard
	Co	59	260.3	1.6	**-0.0033**	0.000	10.0	ug/L	234	Standard
	Ni	60	195.0	3.6	**-0.0341**	0.004	11.2	ug/L	215	Standard
	Cu	65	140.7	10.7	**-0.0385**	0.007	19.0	ug/L	104	Standard
	Zn	66	386.7	3.7	**-0.1493**	0.008	5.7	ug/L	363	Standard
>	**Ge**	72	574185.4	0.9				ug/L	495142	Standard
	As	75	-49.3	11.6	**0.0216**	0.004	19.9	ug/L	-50	Standard
	Se	82	18.5	26.4	**0.0072**	0.039	544.3	ug/L	14	Standard
	Se-1	77	51.3	18.1	**-0.0155**	0.115	739.5	ug/L	42	Standard
>	**Ga**	71	41.7	30.2				mg/L	42	Standard
	Rb	85	33.3	22.9				ug/L	17	Standard
	Y	89	508899.9	2.4				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	287.2	24.5	**0.0857**	0.020	22.8	ug/L	11	Standard
	Ag	107	140.0	24.8	**0.0054**	0.004	76.0	ug/L	68	Standard
	Cd	111	18.8	30.5	**-0.0003**	0.002	828.9	mg/L	2	Standard
	Cd	114	42.6	5.3	**0.0050**	0.000	2.5	ug/L	12	Standard
>	**In**	115	688571.2	3.6				ug/L	587441	Standard
	Sn	118	896.7	13.9	**0.0365**	0.020	55.6	ug/L	497	Standard
	Sb	123	419.7	28.4	**0.0731**	0.025	33.8	ug/L	78	Standard
	Ba	135	58.0	33.3	**0.0081**	0.008	93.7	ug/L	24	Standard
	Ce	140	43.3	17.6				ug/L	25	Standard
>	**Tb**	159	1032903.3	2.1				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	69.3	37.1	**0.0059**	0.003	47.0	ug/L	4	Standard
	Tl	205	78.3	57.2	**0.0101**	0.006	55.2	ug/L	2	Standard
	Pb	206	303.3	6.2	**-0.0195**	0.003	14.1	ug/L	240	Standard
	Pb	207	260.7	8.9	**-0.0200**	0.005	24.1	ug/L	198	Standard
	Pb	208	1157.4	6.9	**-0.0179**	0.003	17.6	ug/L	883	Standard
	U	238	52.0	46.8	**0.0024**	0.003	129.1	ug/L	3	Standard
>	**Bi**	209	575232.2	1.3				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:18:56
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	36.7	28.4	**-0.0241**	0.012	50.4	mg/L	37	Standard
K	39	50.0	26.5	**-0.0190**	0.023	121.8	mg/L	38	Standard
Ca	43	26.7	39.0	**-1.0625**	0.584	55.0	mg/L	10	Standard
Fe	54	147.5	7.0	**-0.0305**	0.005	15.7	mg/L	179	Standard
Fe	57	216.7	8.1	**-0.4394**	0.033	7.6	mg/L	335	Standard
Sc-1	45	70726.5	3.0				mg/L	61029	Standard
Cl	35	29273.6	1.2				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2073.5	10.2				ug/L	1547	Standard
P	31	23289.7	10.5				ug/L	15286	Standard
S	34	2535.2	3.8				ug/L	1572	Standard
Sr	88	68.3	37.5				ug/L	73	Standard
C	12	66.7	22.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	16.3	126.2				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1766.8	31.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.964	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:18:56
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.215
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.780
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:18:56
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138103
Sample Date/Time: Wednesday, June 29, 2016 19:19:52
Number of Replicates: 3
Autosampler Position: 256
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88921.9	2.3				ug/L	74699	Standard
	Be	9	1191.7	0.9	**1.1451**	0.025	2.2	ug/L	8	Standard
	Al	27	19716870.1	2.2	**112.7701**	5.073	4.5	ug/L	1493	Standard
	Sc	45	82006.5	3.2				ug/L	61029	Standard
	Ti	47	8807.2	1.6	**20.0178**	0.269	1.3	ug/L	35	Standard
	V	51	208539.0	1.3	**23.7197**	0.153	0.6	ug/L	1435	Standard
	Cr	52	166877.0	1.1	**18.4795**	0.081	0.4	ug/L	7635	Standard
	Cr	53	19956.5	0.7	**18.3731**	0.126	0.7	ug/L	363	Standard
	Mn	55	4948296.7	1.6	**615.4227**	6.037	1.0	ug/L	747	Standard
	Co	59	82754.8	2.1	**10.8165**	0.154	1.4	ug/L	234	Standard
	Ni	60	51698.7	0.5	**23.7942**	0.298	1.3	ug/L	215	Standard
	Cu	65	48855.0	1.6	**22.6321**	0.340	1.5	ug/L	104	Standard
	Zn	66	105467.2	1.2	**83.1726**	0.384	0.5	ug/L	363	Standard
>	**Ge**	72	571655.8	0.7				ug/L	495142	Standard
	As	75	9636.9	0.9	**7.5116**	0.042	0.6	ug/L	-50	Standard
	Se	82	70.5	8.3	**0.4184**	0.042	10.1	ug/L	14	Standard
	Se-1	77	129.0	9.9	**0.9077**	0.153	16.9	ug/L	42	Standard
>	**Ga**	71	22341.5	2.4				mg/L	42	Standard
	Rb	85	210691.8	1.9				ug/L	17	Standard
	Y	89	776017.6	0.9				ug/L	417098	Standard
>	**Rh**	103	61.7	24.8				ug/L	13	Standard
	Mo	98	1324.4	9.6	**0.4208**	0.045	10.6	ug/L	11	Standard
	Ag	107	918.4	2.4	**0.1080**	0.004	3.7	ug/L	68	Standard
	Cd	111	891.1	4.2	**0.3578**	0.020	5.5	mg/L	2	Standard
	Cd	114	2180.6	6.0	**0.3625**	0.017	4.8	ug/L	12	Standard
>	**In**	115	684819.6	1.2				ug/L	587441	Standard
	Sn	118	1785.1	3.5	**0.1674**	0.006	3.8	ug/L	497	Standard
	Sb	123	354.4	3.1	**0.0609**	0.003	4.6	ug/L	78	Standard
	Ba	135	461342.5	1.4	**193.3434**	2.631	1.4	ug/L	24	Standard
	Ce	140	1368480.3	1.7				ug/L	25	Standard
>	**Tb**	159	1075718.6	2.2				ug/L	939701	Standard
	Ho	165	22026.0	1.7				ug/L	10	Standard
	Tl	203	1135.0	2.0	**0.1247**	0.001	0.6	ug/L	4	Standard
	Tl	205	1093.4	11.6	**0.1404**	0.014	10.2	ug/L	2	Standard
	Pb	206	147882.6	1.3	**25.3549**	0.415	1.6	ug/L	240	Standard
	Pb	207	119862.3	0.6	**23.1268**	0.424	1.8	ug/L	198	Standard
	Pb	208	520445.5	2.0	**25.4242**	0.898	3.5	ug/L	883	Standard
	U	238	4663.1	0.9	**0.6178**	0.006	0.9	ug/L	3	Standard
>	**Bi**	209	570570.9	1.5				ug/L	523984	Standard

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:22:09
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	126.7	27.7	**0.0563**	0.032	56.3	mg/L	37	Standard
K	39	878.4	3.1	**1.3326**	0.091	6.9	mg/L	38	Standard
Ca	43	426.7	4.7	**17.5293**	0.460	2.6	mg/L	10	Standard
Fe	54	21094.2	3.4	**13.1896**	0.865	6.6	mg/L	179	Standard
Fe	57	5826.1	1.4	**13.6721**	0.491	3.6	mg/L	335	Standard
> **Sc-1**	45	82006.5	3.2				mg/L	61029	Standard
Cl	35	28772.6	2.0				ug/L	20195	Standard
Kr	83	3.0	66.7				ug/L	4	Standard
Br	81	3580.4	3.6				ug/L	1547	Standard
P	31	21750.6	1.3				ug/L	15286	Standard
S	34	2703.6	0.6				ug/L	1572	Standard
Sr	88	178.3	13.8				ug/L	73	Standard
C	12	233.3	9.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	96.7	36.3				mg/L	3	Standard
Dy	164	37298.5	1.0				mg/L	6	Standard
Ho-1	165	22026.0	1.7				mg/L	10	Standard
Er	166	20383.7	1.3				mg/L	13	Standard
I	127	19701.1	3.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		119.041	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.453	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:22:09
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	116.577
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.891
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:22:09
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138103
Sample Date/Time: Wednesday, June 29, 2016 19:23:03
Number of Replicates: 3
Autosampler Position: 257
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82351.7	3.3				ug/L	74699	Standard
	Be	9	43.3	24.0	**0.0523**	0.011	20.3	ug/L	8	Standard
	Al	27	430174.7	3.0	**2.6419**	0.016	0.6	ug/L	1493	Standard
	Sc	45	73040.6	1.2				ug/L	61029	Standard
	Ti	47	215.7	6.2	**0.3839**	0.033	8.6	ug/L	35	Standard
	V	51	5937.3	2.7	**0.4853**	0.017	3.5	ug/L	1435	Standard
	Cr	52	11052.0	1.3	**0.2679**	0.017	6.2	ug/L	7635	Standard
	Cr	53	723.4	5.6	**0.3480**	0.038	11.0	ug/L	363	Standard
	Mn	55	109527.8	0.6	**13.4774**	0.102	0.8	ug/L	747	Standard
	Co	59	1998.1	1.8	**0.2248**	0.006	2.5	ug/L	234	Standard
	Ni	60	1283.7	2.7	**0.4700**	0.016	3.4	ug/L	215	Standard
	Cu	65	1177.7	1.9	**0.4444**	0.009	2.1	ug/L	104	Standard
	Zn	66	4258.3	1.3	**2.9218**	0.029	1.0	ug/L	363	Standard
>	**Ge**	72	571680.5	0.6				ug/L	495142	Standard
	As	75	130.4	5.6	**0.1604**	0.006	3.7	ug/L	-50	Standard
	Se	82	14.1	16.1	**-0.0268**	0.018	65.5	ug/L	14	Standard
	Se-1	77	58.0	6.9	**0.0653**	0.044	66.8	ug/L	42	Standard
>	**Ga**	71	521.7	4.0				mg/L	42	Standard
	Rb	85	4554.0	2.0				ug/L	17	Standard
	Y	89	512515.9	2.6				ug/L	417098	Standard
>	**Rh**	103	11.7	24.7				ug/L	13	Standard
	Mo	98	90.4	4.4	**0.0229**	0.001	6.5	ug/L	11	Standard
	Ag	107	144.3	4.6	**0.0058**	0.001	16.1	ug/L	68	Standard
	Cd	111	21.8	34.6	**0.0009**	0.003	335.6	mg/L	2	Standard
	Cd	114	67.8	46.0	**0.0090**	0.005	55.6	ug/L	12	Standard
>	**In**	115	696556.9	1.5				ug/L	587441	Standard
	Sn	118	640.0	3.6	**-0.0023**	0.005	204.9	ug/L	497	Standard
	Sb	123	182.0	15.1	**0.0287**	0.006	19.2	ug/L	78	Standard
	Ba	135	9992.3	2.7	**4.1009**	0.069	1.7	ug/L	24	Standard
	Ce	140	29587.2	2.0				ug/L	25	Standard
>	**Tb**	159	1058949.6	2.5				ug/L	939701	Standard
	Ho	165	493.3	6.5				ug/L	10	Standard
	Tl	203	41.7	21.8	**0.0028**	0.001	37.1	ug/L	4	Standard
	Tl	205	43.3	54.5	**0.0057**	0.003	52.4	ug/L	2	Standard
	Pb	206	3583.8	1.2	**0.5358**	0.012	2.2	ug/L	240	Standard
	Pb	207	2855.9	1.4	**0.4744**	0.011	2.4	ug/L	198	Standard
	Pb	208	12510.0	0.4	**0.5295**	0.006	1.1	ug/L	883	Standard
	U	238	110.3	6.2	**0.0101**	0.001	8.0	ug/L	3	Standard
>	**Bi**	209	579224.4	1.1				ug/L	523984	Standard

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:25:20
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	51.7	29.6	**-0.0092**	0.017	184.4	mg/L	37	Standard
K	39	41.7	61.6	**-0.0375**	0.047	124.3	mg/L	38	Standard
Ca	43	26.7	10.8	**-1.1162**	0.146	13.1	mg/L	10	Standard
Fe	54	584.0	15.6	**0.2752**	0.062	22.7	mg/L	179	Standard
Fe	57	388.3	5.2	**0.0289**	0.063	219.5	mg/L	335	Standard
> **Sc-1**	45	73040.6	1.2				mg/L	61029	Standard
Cl	35	28933.6	1.8				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2263.5	5.9				ug/L	1547	Standard
P	31	20116.7	3.4				ug/L	15286	Standard
S	34	2530.2	7.0				ug/L	1572	Standard
Sr	88	75.0	23.1				ug/L	73	Standard
C	12	66.7	56.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	963.7	4.1				mg/L	6	Standard
Ho-1	165	493.3	6.5				mg/L	10	Standard
Er	166	410.0	6.5				mg/L	13	Standard
I	127	2373.5	31.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.245	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.458	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:25:20
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		118.575
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.542
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138104
Sample Date/Time: Wednesday, June 29, 2016 19:26:15
Number of Replicates: 3
Autosampler Position: 258
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85261.0	2.3				ug/L	74699	Standard
	Be	9	1065.0	16.2	**1.0661**	0.158	14.8	ug/L	8	Standard
	Al	27	15142099.3	1.2	**90.2750**	1.228	1.4	ug/L	1493	Standard
	Sc	45	77933.3	2.2				ug/L	61029	Standard
	Ti	47	7118.0	1.5	**16.2707**	0.118	0.7	ug/L	35	Standard
	V	51	161044.2	0.8	**18.4034**	0.196	1.1	ug/L	1435	Standard
	Cr	52	146045.2	0.2	**16.1669**	0.325	2.0	ug/L	7635	Standard
	Cr	53	17793.8	1.2	**16.4664**	0.462	2.8	ug/L	363	Standard
	Mn	55	4299510.8	1.0	**538.5113**	8.104	1.5	ug/L	747	Standard
	Co	59	67815.8	1.0	**8.9211**	0.182	2.0	ug/L	234	Standard
	Ni	60	46536.4	1.1	**21.5551**	0.152	0.7	ug/L	215	Standard
	Cu	65	28919.6	1.2	**13.4506**	0.294	2.2	ug/L	104	Standard
	Zn	66	52781.1	1.6	**41.6917**	0.739	1.8	ug/L	363	Standard
>	**Ge**	72	567729.5	1.7				ug/L	495142	Standard
	As	75	4550.5	2.3	**3.6043**	0.141	3.9	ug/L	-50	Standard
	Se	82	43.5	41.5	**0.2091**	0.148	70.8	ug/L	14	Standard
	Se-1	77	118.3	4.0	**0.7906**	0.045	5.7	ug/L	42	Standard
>	**Ga**	71	21335.0	2.6				mg/L	42	Standard
	Rb	85	199689.0	1.1				ug/L	17	Standard
	Y	89	717025.3	1.7				ug/L	417098	Standard
>	**Rh**	103	53.3	39.0				ug/L	13	Standard
	Mo	98	750.2	3.3	**0.2455**	0.003	1.3	ug/L	11	Standard
	Ag	107	572.7	8.1	**0.0655**	0.005	7.5	ug/L	68	Standard
	Cd	111	313.3	5.6	**0.1259**	0.007	5.6	mg/L	2	Standard
	Cd	114	729.6	4.4	**0.1250**	0.007	5.3	ug/L	12	Standard
>	**In**	115	657920.1	2.1				ug/L	587441	Standard
	Sn	118	513.3	6.6	**-0.0162**	0.005	32.2	ug/L	497	Standard
	Sb	123	156.7	10.3	**0.0258**	0.004	14.3	ug/L	78	Standard
	Ba	135	278778.0	0.4	**121.6374**	2.867	2.4	ug/L	24	Standard
	Ce	140	1205012.6	0.9				ug/L	25	Standard
>	**Tb**	159	1028655.6	1.0				ug/L	939701	Standard
	Ho	165	19245.5	0.5				ug/L	10	Standard
	Tl	203	1022.7	4.2	**0.1168**	0.004	3.6	ug/L	4	Standard
	Tl	205	930.0	6.2	**0.1244**	0.008	6.1	ug/L	2	Standard
	Pb	206	75172.7	1.1	**13.3810**	0.239	1.8	ug/L	240	Standard
	Pb	207	60753.2	0.6	**12.1662**	0.152	1.3	ug/L	198	Standard
	Pb	208	268421.8	0.9	**13.6089**	0.053	0.4	ug/L	883	Standard
	U	238	4956.8	1.6	**0.6840**	0.014	2.1	ug/L	3	Standard
>	**Bi**	209	548179.9	0.7				ug/L	523984	Standard

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:28:31
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**29.7760**	51.565	173.2	mg/L	0	Standard
Mg	24	116.7	8.9	**0.0527**	0.008	15.2	mg/L	37	Standard
K	39	606.7	4.1	**0.9368**	0.065	7.0	mg/L	38	Standard
Ca	43	223.3	26.5	**8.5548**	3.074	35.9	mg/L	10	Standard
Fe	54	18758.0	3.4	**12.3136**	0.167	1.4	mg/L	179	Standard
Fe	57	5059.2	5.4	**12.3881**	0.534	4.3	mg/L	335	Standard
Sc-1	45	77933.3	2.2				mg/L	61029	Standard
Cl	35	27511.6	2.5				ug/L	20195	Standard
Kr	83	4.0	50.0				ug/L	4	Standard
Br	81	2450.2	6.4				ug/L	1547	Standard
P	31	19838.0	2.4				ug/L	15286	Standard
S	34	2438.5	4.0				ug/L	1572	Standard
Sr	88	130.0	10.2				ug/L	73	Standard
C	12	76.7	41.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	16.7	124.9				mg/L	3	Standard
Dy	164	31319.3	1.6				mg/L	6	Standard
Ho-1	165	19245.5	0.5				mg/L	10	Standard
Er	166	17009.5	2.0				mg/L	13	Standard
I	127	14255.0	0.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		114.140	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.660	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:28:31
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	111.998
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.618
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:28:31
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138104
Sample Date/Time: Wednesday, June 29, 2016 19:29:26
Number of Replicates: 3
Autosampler Position: 259
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81019.4	0.7				ug/L	74699	Standard
	Be	9	41.7	18.3	**0.0512**	0.008	15.3	ug/L	8	Standard
	Al	27	329954.8	1.9	**2.0568**	0.044	2.2	ug/L	1493	Standard
	Sc	45	70853.8	4.4				ug/L	61029	Standard
	Ti	47	170.0	4.8	**0.2650**	0.022	8.2	ug/L	35	Standard
	V	51	4919.9	1.5	**0.3475**	0.009	2.6	ug/L	1435	Standard
	Cr	52	11144.8	0.5	**0.2299**	0.010	4.3	ug/L	7635	Standard
	Cr	53	798.4	9.1	**0.3907**	0.073	18.6	ug/L	363	Standard
	Mn	55	93006.7	2.2	**10.9903**	0.351	3.2	ug/L	747	Standard
	Co	59	1669.1	4.9	**0.1734**	0.010	5.5	ug/L	234	Standard
	Ni	60	1239.4	3.4	**0.4280**	0.019	4.5	ug/L	215	Standard
	Cu	65	783.4	0.7	**0.2472**	0.006	2.3	ug/L	104	Standard
	Zn	66	3223.7	4.0	**2.0065**	0.120	6.0	ug/L	363	Standard
>	**Ge**	72	593953.7	1.0				ug/L	495142	Standard
	As	75	25.6	116.2	**0.0786**	0.022	28.0	ug/L	-50	Standard
	Se	82	15.5	11.3	**-0.0201**	0.014	71.9	ug/L	14	Standard
	Se-1	77	55.7	3.7	**0.0130**	0.019	145.3	ug/L	42	Standard
>	**Ga**	71	496.7	9.6				mg/L	42	Standard
	Rb	85	4318.9	1.6				ug/L	17	Standard
	Y	89	493099.4	1.2				ug/L	417098	Standard
>	**Rh**	103	8.3	91.7				ug/L	13	Standard
	Mo	98	63.8	10.3	**0.0154**	0.002	13.8	ug/L	11	Standard
	Ag	107	130.3	15.7	**0.0047**	0.003	56.8	ug/L	68	Standard
	Cd	111	7.2	21.0	**-0.0049**	0.001	12.8	mg/L	2	Standard
	Cd	114	27.0	44.7	**0.0025**	0.002	80.8	ug/L	12	Standard
>	**In**	115	665068.2	0.5				ug/L	587441	Standard
	Sn	118	545.0	15.7	**-0.0123**	0.013	107.4	ug/L	497	Standard
	Sb	123	150.6	9.1	**0.0243**	0.003	11.0	ug/L	78	Standard
	Ba	135	6015.5	0.5	**2.5800**	0.016	0.6	ug/L	24	Standard
	Ce	140	25678.4	3.6				ug/L	25	Standard
>	**Tb**	159	1023204.1	0.9				ug/L	939701	Standard
	Ho	165	420.0	5.5				ug/L	10	Standard
	Tl	203	37.3	22.8	**0.0024**	0.001	40.1	ug/L	4	Standard
	Tl	205	30.0	50.0	**0.0040**	0.002	46.9	ug/L	2	Standard
	Pb	206	1995.8	4.0	**0.2734**	0.016	6.0	ug/L	240	Standard
	Pb	207	1576.4	5.4	**0.2364**	0.019	8.0	ug/L	198	Standard
	Pb	208	6805.9	1.7	**0.2606**	0.003	1.1	ug/L	883	Standard
	U	238	99.3	2.5	**0.0089**	0.000	2.4	ug/L	3	Standard
>	**Bi**	209	568219.3	0.9				ug/L	523984	Standard

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:31:43
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	30.0	66.7	**-0.0312**	0.023	74.2	mg/L	37	Standard
K	39	65.0	30.8	**0.0108**	0.043	395.4	mg/L	38	Standard
Ca	43	20.0	0.0	**-1.4338**	0.047	3.3	mg/L	10	Standard
Fe	54	499.3	4.0	**0.2265**	0.010	4.2	mg/L	179	Standard
Fe	57	346.7	6.8	**-0.0592**	0.056	94.8	mg/L	335	Standard
> **Sc-1**	45	70853.8	4.4				mg/L	61029	Standard
Cl	35	26917.9	3.4				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	1996.8	2.8				ug/L	1547	Standard
P	31	19530.9	2.5				ug/L	15286	Standard
S	34	2235.2	2.4				ug/L	1572	Standard
Sr	88	70.0	18.9				ug/L	73	Standard
C	12	83.3	38.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	818.0	9.1				mg/L	6	Standard
Ho-1	165	420.0	5.5				mg/L	10	Standard
Er	166	390.0	11.8				mg/L	13	Standard
I	127	1930.1	32.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		108.461	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		119.956	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:31:43
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.214
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.442
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138105
Sample Date/Time: Wednesday, June 29, 2016 19:32:37
Number of Replicates: 3
Autosampler Position: 260
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86630.3	1.6				ug/L	74699	Standard
	Be	9	1345.1	15.8	**1.3230**	0.188	14.2	ug/L	8	Standard
	Al	27	25518871.0	0.5	**149.7364**	2.025	1.4	ug/L	1493	Standard
	Sc	45	82793.9	1.1				ug/L	61029	Standard
	Ti	47	9047.0	1.1	**21.0343**	0.216	1.0	ug/L	35	Standard
	V	51	241144.8	2.1	**28.0828**	0.136	0.5	ug/L	1435	Standard
	Cr	52	185820.6	3.2	**21.1812**	0.417	2.0	ug/L	7635	Standard
	Cr	53	22667.0	3.7	**21.3886**	0.515	2.4	ug/L	363	Standard
	Mn	55	5666490.2	1.7	**720.7034**	2.244	0.3	ug/L	747	Standard
	Co	59	89984.1	1.6	**12.0322**	0.097	0.8	ug/L	234	Standard
	Ni	60	57928.9	2.2	**27.2782**	0.178	0.7	ug/L	215	Standard
	Cu	65	41127.1	3.0	**19.4650**	0.281	1.4	ug/L	104	Standard
	Zn	66	70831.6	3.1	**56.9699**	0.880	1.5	ug/L	363	Standard
>	**Ge**	72	559032.2	1.6				ug/L	495142	Standard
	As	75	9088.7	3.2	**7.2452**	0.123	1.7	ug/L	-50	Standard
	Se	82	66.3	23.6	**0.3960**	0.117	29.7	ug/L	14	Standard
	Se-1	77	130.0	4.3	**0.9538**	0.045	4.7	ug/L	42	Standard
>	**Ga**	71	25715.1	2.4				mg/L	42	Standard
	Rb	85	243499.9	2.1				ug/L	17	Standard
	Y	89	766854.2	0.5				ug/L	417098	Standard
>	**Rh**	103	51.7	5.6				ug/L	13	Standard
	Mo	98	760.8	5.6	**0.2496**	0.019	7.7	ug/L	11	Standard
	Ag	107	727.0	8.3	**0.0869**	0.010	11.3	ug/L	68	Standard
	Cd	111	420.9	5.2	**0.1721**	0.013	7.6	mg/L	2	Standard
	Cd	114	1021.1	15.6	**0.1761**	0.030	17.3	ug/L	12	Standard
>	**In**	115	657583.1	2.2				ug/L	587441	Standard
	Sn	118	706.7	9.4	**0.0135**	0.012	91.4	ug/L	497	Standard
	Sb	123	160.0	3.9	**0.0265**	0.002	7.0	ug/L	78	Standard
	Ba	135	489546.4	1.0	**213.7510**	6.703	3.1	ug/L	24	Standard
	Ce	140	1473172.3	0.9				ug/L	25	Standard
>	**Tb**	159	1043758.3	1.4				ug/L	939701	Standard
	Ho	165	23187.7	0.9				ug/L	10	Standard
	Tl	203	1287.4	7.3	**0.1458**	0.012	8.5	ug/L	4	Standard
	Tl	205	1168.4	5.5	**0.1545**	0.010	6.8	ug/L	2	Standard
	Pb	206	110607.4	2.3	**19.4839**	0.695	3.6	ug/L	240	Standard
	Pb	207	89635.8	2.2	**17.7667**	0.631	3.6	ug/L	198	Standard
	Pb	208	392577.0	0.9	**19.6965**	0.434	2.2	ug/L	883	Standard
	U	238	8119.8	0.8	**1.1095**	0.020	1.8	ug/L	3	Standard
>	**Bi**	209	554986.0	1.4				ug/L	523984	Standard

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:34:54
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**42.5482**	42.299	99.4	mg/L	0	Standard
Mg	24	165.0	16.0	**0.0921**	0.027	29.3	mg/L	37	Standard
K	39	870.0	2.5	**1.3032**	0.021	1.6	mg/L	38	Standard
Ca	43	623.3	6.5	**26.5121**	2.122	8.0	mg/L	10	Standard
Fe	54	24836.5	3.4	**15.3836**	0.500	3.3	mg/L	179	Standard
Fe	57	6766.5	3.6	**15.8771**	0.513	3.2	mg/L	335	Standard
> **Sc-1**	45	82793.9	1.1				mg/L	61029	Standard
Cl	35	27591.1	0.8				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2510.2	7.0				ug/L	1547	Standard
P	31	20592.3	2.0				ug/L	15286	Standard
S	34	2640.2	1.7				ug/L	1572	Standard
Sr	88	220.0	6.0				ug/L	73	Standard
C	12	86.7	24.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	46.7	44.6				mg/L	3	Standard
Dy	164	40115.6	1.1				mg/L	6	Standard
Ho-1	165	23187.7	0.9				mg/L	10	Standard
Er	166	21645.5	1.0				mg/L	13	Standard
I	127	13629.5	5.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		115.973	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		112.903	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:34:54
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	111.940
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	105.916
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:34:54
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138105
Sample Date/Time: Wednesday, June 29, 2016 19:35:50
Number of Replicates: 3
Autosampler Position: 401
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81775.3	3.6				ug/L	74699	Standard
	Be	9	45.0	11.1	**0.0544**	0.006	11.5	ug/L	8	Standard
	Al	27	547945.2	2.8	**3.3933**	0.048	1.4	ug/L	1493	Standard
	Sc	45	73444.3	1.5				ug/L	61029	Standard
	Ti	47	212.7	8.9	**0.3795**	0.037	9.9	ug/L	35	Standard
	V	51	6482.3	2.1	**0.5523**	0.015	2.6	ug/L	1435	Standard
	Cr	52	11556.7	2.9	**0.3347**	0.014	4.0	ug/L	7635	Standard
	Cr	53	865.0	10.6	**0.4848**	0.072	14.9	ug/L	363	Standard
	Mn	55	123099.8	1.9	**15.2575**	0.218	1.4	ug/L	747	Standard
	Co	59	2101.5	1.5	**0.2401**	0.008	3.5	ug/L	234	Standard
	Ni	60	1463.7	4.5	**0.5572**	0.024	4.3	ug/L	215	Standard
	Cu	65	1004.0	2.5	**0.3664**	0.012	3.3	ug/L	104	Standard
	Zn	66	6132.9	2.2	**4.4370**	0.037	0.8	ug/L	363	Standard
>	**Ge**	72	568314.4	1.9				ug/L	495142	Standard
	As	75	135.4	21.1	**0.1651**	0.024	14.5	ug/L	-50	Standard
	Se	82	20.1	25.4	**0.0217**	0.043	197.6	ug/L	14	Standard
	Se-1	77	47.3	2.4	**-0.0575**	0.019	32.3	ug/L	42	Standard
>	**Ga**	71	615.0	5.6				mg/L	42	Standard
	Rb	85	5134.2	2.0				ug/L	17	Standard
	Y	89	506159.2	1.9				ug/L	417098	Standard
>	**Rh**	103	11.7	107.9				ug/L	13	Standard
	Mo	98	50.7	18.0	**0.0105**	0.003	28.3	ug/L	11	Standard
	Ag	107	125.7	10.6	**0.0035**	0.002	48.0	ug/L	68	Standard
	Cd	111	10.9	48.5	**-0.0035**	0.002	62.7	mg/L	2	Standard
	Cd	114	42.4	34.4	**0.0049**	0.002	48.6	ug/L	12	Standard
>	**In**	115	689821.8	0.4				ug/L	587441	Standard
	Sn	118	595.0	5.8	**-0.0080**	0.005	59.8	ug/L	497	Standard
	Sb	123	118.9	14.7	**0.0175**	0.003	18.8	ug/L	78	Standard
	Ba	135	10423.6	1.2	**4.3208**	0.035	0.8	ug/L	24	Standard
	Ce	140	31651.4	0.9				ug/L	25	Standard
>	**Tb**	159	1043112.5	1.4				ug/L	939701	Standard
	Ho	165	500.0	7.2				ug/L	10	Standard
	Tl	203	34.0	12.8	**0.0020**	0.000	23.2	ug/L	4	Standard
	Tl	205	20.0	66.1	**0.0027**	0.002	61.7	ug/L	2	Standard
	Pb	206	2645.2	1.4	**0.3784**	0.005	1.3	ug/L	240	Standard
	Pb	207	2160.5	2.6	**0.3433**	0.011	3.1	ug/L	198	Standard
	Pb	208	9496.8	0.4	**0.3858**	0.004	1.1	ug/L	883	Standard
	U	238	183.7	2.7	**0.0198**	0.001	4.5	ug/L	3	Standard
>	**Bi**	209	577076.9	1.0				ug/L	523984	Standard

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:38:07
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**15.9398**	27.600	173.2	mg/L	0	Standard
Mg	24	31.7	39.7	**-0.0310**	0.014	45.0	mg/L	37	Standard
K	39	75.0	17.6	**0.0236**	0.024	101.0	mg/L	38	Standard
Ca	43	45.0	44.4	**-0.1502**	1.087	724.0	mg/L	10	Standard
Fe	54	679.1	0.8	**0.3402**	0.010	2.9	mg/L	179	Standard
Fe	57	390.0	9.7	**0.0271**	0.102	376.6	mg/L	335	Standard
Sc-1	45	73444.3	1.5				mg/L	61029	Standard
Cl	35	27695.3	0.4				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	2046.8	15.0				ug/L	1547	Standard
P	31	20195.1	3.7				ug/L	15286	Standard
S	34	2511.9	7.3				ug/L	1572	Standard
Sr	88	68.3	42.9				ug/L	73	Standard
C	12	50.0	40.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	1042.0	9.5				mg/L	6	Standard
Ho-1	165	500.0	7.2				mg/L	10	Standard
Er	166	446.7	15.2				mg/L	13	Standard
I	127	1841.8	32.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		109.473	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.778	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:38:07
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.428
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.132
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138107
Sample Date/Time: Wednesday, June 29, 2016 19:39:01
Number of Replicates: 3
Autosampler Position: 402
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84183.2	0.7				ug/L	74699	Standard
	Be	9	1125.0	5.6	**1.1418**	0.071	6.2	ug/L	8	Standard
	Al	27	16362118.1	1.0	**98.7783**	0.357	0.4	ug/L	1493	Standard
	Sc	45	80079.5	2.6				ug/L	61029	Standard
	Ti	47	5684.7	1.2	**13.0248**	0.178	1.4	ug/L	35	Standard
	V	51	193164.5	0.6	**22.2022**	0.395	1.8	ug/L	1435	Standard
	Cr	52	144926.2	1.3	**16.0993**	0.107	0.7	ug/L	7635	Standard
	Cr	53	17089.6	2.1	**15.8598**	0.165	1.0	ug/L	363	Standard
	Mn	55	4945150.4	1.2	**621.8545**	12.591	2.0	ug/L	747	Standard
	Co	59	79442.1	0.4	**10.4972**	0.101	1.0	ug/L	234	Standard
	Ni	60	48759.0	0.6	**22.6802**	0.156	0.7	ug/L	215	Standard
	Cu	65	44548.4	1.0	**20.8546**	0.085	0.4	ug/L	104	Standard
	Zn	66	116987.3	1.0	**93.3289**	1.249	1.3	ug/L	363	Standard
>	**Ge**	72	565487.8	1.3				ug/L	495142	Standard
	As	75	12786.6	1.5	**10.0557**	0.155	1.5	ug/L	-50	Standard
	Se	82	93.9	6.5	**0.6112**	0.058	9.6	ug/L	14	Standard
	Se-1	77	134.7	12.3	**0.9941**	0.220	22.1	ug/L	42	Standard
>	**Ga**	71	20750.9	1.1				mg/L	42	Standard
	Rb	85	229576.4	1.2				ug/L	17	Standard
	Y	89	760564.3	0.9				ug/L	417098	Standard
>	**Rh**	103	26.7	39.0				ug/L	13	Standard
	Mo	98	714.7	6.9	**0.2279**	0.017	7.3	ug/L	11	Standard
	Ag	107	706.7	2.8	**0.0816**	0.003	4.3	ug/L	68	Standard
	Cd	111	991.8	3.3	**0.4053**	0.015	3.7	mg/L	2	Standard
	Cd	114	2256.1	2.3	**0.3810**	0.005	1.2	ug/L	12	Standard
>	**In**	115	674434.0	1.0				ug/L	587441	Standard
	Sn	118	836.7	10.4	**0.0300**	0.012	41.1	ug/L	497	Standard
	Sb	123	542.0	6.0	**0.0969**	0.005	5.6	ug/L	78	Standard
	Ba	135	363437.5	1.5	**154.6607**	2.916	1.9	ug/L	24	Standard
	Ce	140	1323769.8	0.4				ug/L	25	Standard
>	**Tb**	159	1046974.9	0.8				ug/L	939701	Standard
	Ho	165	21420.1	0.8				ug/L	10	Standard
	Tl	203	1015.7	4.2	**0.1138**	0.005	4.5	ug/L	4	Standard
	Tl	205	906.7	2.2	**0.1191**	0.002	2.0	ug/L	2	Standard
	Pb	206	225485.9	0.9	**39.5300**	0.272	0.7	ug/L	240	Standard
	Pb	207	184080.5	0.2	**36.3178**	0.069	0.2	ug/L	198	Standard
	Pb	208	811320.1	0.8	**40.5199**	0.434	1.1	ug/L	883	Standard
	U	238	5280.9	2.0	**0.7154**	0.017	2.3	ug/L	3	Standard
>	**Bi**	209	558508.9	0.3				ug/L	523984	Standard

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:41:18
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	105.0	49.7	**0.0390**	0.054	139.2	mg/L	37	Standard
K	39	875.0	3.0	**1.3597**	0.017	1.3	mg/L	38	Standard
Ca	43	123.3	10.2	**3.4107**	0.543	15.9	mg/L	10	Standard
Fe	54	18540.3	2.4	**11.8486**	0.489	4.1	mg/L	179	Standard
Fe	57	5050.8	2.1	**12.0198**	0.614	5.1	mg/L	335	Standard
> **Sc-1**	45	80079.5	2.6				mg/L	61029	Standard
Cl	35	26211.3	2.3				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	4010.5	1.9				ug/L	1547	Standard
P	31	18821.7	2.8				ug/L	15286	Standard
S	34	2180.2	5.1				ug/L	1572	Standard
Sr	88	100.0	25.0				ug/L	73	Standard
C	12	166.7	30.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	70.0	28.6				mg/L	3	Standard
Dy	164	35314.8	2.4				mg/L	6	Standard
Ho-1	165	21420.1	0.8				mg/L	10	Standard
Er	166	18965.2	4.5				mg/L	13	Standard
I	127	21211.5	3.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		112.697	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.207	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:41:18
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	114.809
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	106.589
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:41:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138107
Sample Date/Time: Wednesday, June 29, 2016 19:42:13
Number of Replicates: 3
Autosampler Position: 403
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79962.1	1.1				ug/L	74699	Standard
	Be	9	35.0	37.8	**0.0449**	0.015	32.4	ug/L	8	Standard
	Al	27	348986.4	2.3	**2.2050**	0.043	1.9	ug/L	1493	Standard
	Sc	45	69050.4	2.1				ug/L	61029	Standard
	Ti	47	147.7	4.0	**0.2190**	0.009	4.2	ug/L	35	Standard
	V	51	5537.0	5.1	**0.4222**	0.040	9.5	ug/L	1435	Standard
	Cr	52	11219.2	0.9	**0.2513**	0.015	5.8	ug/L	7635	Standard
	Cr	53	801.7	3.8	**0.4006**	0.025	6.2	ug/L	363	Standard
	Mn	55	101661.8	3.0	**12.1540**	0.524	4.3	ug/L	747	Standard
	Co	59	1813.1	5.5	**0.1941**	0.015	7.9	ug/L	234	Standard
	Ni	60	1251.7	0.2	**0.4392**	0.009	2.1	ug/L	215	Standard
	Cu	65	1098.7	4.3	**0.3937**	0.026	6.7	ug/L	104	Standard
	Zn	66	4663.4	1.9	**3.1415**	0.081	2.6	ug/L	363	Standard
>	**Ge**	72	587910.1	1.4				ug/L	495142	Standard
	As	75	208.3	5.8	**0.2162**	0.008	3.9	ug/L	-50	Standard
	Se	82	14.6	52.1	**-0.0264**	0.056	213.9	ug/L	14	Standard
	Se-1	77	58.3	15.6	**0.0509**	0.111	217.0	ug/L	42	Standard
>	**Ga**	71	483.3	11.4				mg/L	42	Standard
	Rb	85	4672.4	3.2				ug/L	17	Standard
	Y	89	494582.5	2.5				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	50.9	2.1	**0.0113**	0.001	5.2	ug/L	11	Standard
	Ag	107	137.0	5.3	**0.0059**	0.001	14.4	ug/L	68	Standard
	Cd	111	22.2	22.2	**0.0016**	0.002	123.1	mg/L	2	Standard
	Cd	114	55.9	25.6	**0.0077**	0.003	34.0	ug/L	12	Standard
>	**In**	115	657881.3	1.7				ug/L	587441	Standard
	Sn	118	586.7	15.6	**-0.0052**	0.012	239.2	ug/L	497	Standard
	Sb	123	130.5	4.0	**0.0208**	0.001	5.8	ug/L	78	Standard
	Ba	135	7584.6	1.4	**3.2933**	0.054	1.6	ug/L	24	Standard
	Ce	140	27601.8	1.0				ug/L	25	Standard
>	**Tb**	159	1004279.0	0.4				ug/L	939701	Standard
	Ho	165	470.0	12.8				ug/L	10	Standard
	Tl	203	31.3	21.7	**0.0018**	0.001	42.5	ug/L	4	Standard
	Tl	205	38.3	39.8	**0.0052**	0.002	37.2	ug/L	2	Standard
	Pb	206	5039.5	2.9	**0.8074**	0.036	4.5	ug/L	240	Standard
	Pb	207	4082.6	3.5	**0.7312**	0.039	5.3	ug/L	198	Standard
	Pb	208	17367.6	1.6	**0.7884**	0.025	3.2	ug/L	883	Standard
	U	238	118.3	14.2	**0.0116**	0.002	21.2	ug/L	3	Standard
>	**Bi**	209	562827.5	1.3				ug/L	523984	Standard

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:44:30
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	41.7	6.9	**-0.0175**	0.004	24.7	mg/L	37	Standard
K	39	65.0	46.8	**0.0126**	0.058	464.1	mg/L	38	Standard
Ca	43	20.0	66.1	**-1.4118**	0.719	50.9	mg/L	10	Standard
Fe	54	548.6	17.9	**0.2728**	0.072	26.3	mg/L	179	Standard
Fe	57	318.3	8.7	**-0.1165**	0.103	88.6	mg/L	335	Standard
> **Sc-1**	45	69050.4	2.1				mg/L	61029	Standard
Cl	35	24683.4	2.1				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	2036.8	6.4				ug/L	1547	Standard
P	31	17868.9	0.9				ug/L	15286	Standard
S	34	1880.1	5.1				ug/L	1572	Standard
Sr	88	70.0	32.7				ug/L	73	Standard
C	12	63.3	32.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	851.3	21.1				mg/L	6	Standard
Ho-1	165	470.0	12.8				mg/L	10	Standard
Er	166	460.0	19.0				mg/L	13	Standard
I	127	2083.5	32.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		107.046	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		118.736	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:44:30
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.991
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.413
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138402
Sample Date/Time: Wednesday, June 29, 2016 19:45:24
Number of Replicates: 3
Autosampler Position: 404
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78829.6	2.0				ug/L	74699	Standard
	Be	9	476.7	4.8	**0.5211**	0.034	6.6	ug/L	8	Standard
	Al	27	790696.4	0.9	**5.0858**	0.055	1.1	ug/L	1493	Standard
	Sc	45	79546.6	1.7				ug/L	61029	Standard
	Ti	47	5657.7	0.8	**12.8152**	0.349	2.7	ug/L	35	Standard
	V	51	106854.5	1.6	**12.0511**	0.193	1.6	ug/L	1435	Standard
	Cr	52	263051.4	1.4	**29.7045**	0.729	2.5	ug/L	7635	Standard
	Cr	53	32229.3	4.0	**29.8443**	0.696	2.3	ug/L	363	Standard
	Mn	55	1259215.6	0.6	**156.4341**	3.728	2.4	ug/L	747	Standard
	Co	59	45647.0	0.6	**5.9475**	0.157	2.6	ug/L	234	Standard
	Ni	60	134444.3	1.6	**62.0317**	0.857	1.4	ug/L	215	Standard
	Cu	65	50137.5	0.5	**23.2180**	0.523	2.3	ug/L	104	Standard
	Zn	66	191528.7	0.9	**151.3328**	2.746	1.8	ug/L	363	Standard
>	**Ge**	72	572152.6	2.7				ug/L	495142	Standard
	As	75	5700.3	1.6	**4.4645**	0.048	1.1	ug/L	-50	Standard
	Se	82	330.0	4.2	**2.4632**	0.049	2.0	ug/L	14	Standard
	Se-1	77	314.7	3.0	**3.1071**	0.115	3.7	ug/L	42	Standard
>	**Ga**	71	3685.4	1.6				mg/L	42	Standard
	Rb	85	41053.9	1.7				ug/L	17	Standard
	Y	89	1388727.0	1.6				ug/L	417098	Standard
>	**Rh**	103	25.0	0.0				ug/L	13	Standard
	Mo	98	3897.2	2.0	**1.2782**	0.019	1.5	ug/L	11	Standard
	Ag	107	551.3	5.8	**0.0614**	0.003	5.7	ug/L	68	Standard
	Cd	111	1240.5	0.7	**0.5131**	0.007	1.4	mg/L	2	Standard
	Cd	114	2915.5	4.4	**0.4970**	0.020	4.1	ug/L	12	Standard
>	**In**	115	669008.5	1.2				ug/L	587441	Standard
	Sn	118	676.7	8.2	**0.0070**	0.008	114.7	ug/L	497	Standard
	Sb	123	186.6	3.8	**0.0309**	0.001	3.8	ug/L	78	Standard
	Ba	135	221251.7	0.3	**94.9139**	1.345	1.4	ug/L	24	Standard
	Ce	140	471655.7	0.9				ug/L	25	Standard
>	**Tb**	159	1067869.0	0.9				ug/L	939701	Standard
	Ho	165	47698.3	1.5				ug/L	10	Standard
	Tl	203	1397.7	2.5	**0.1543**	0.005	3.4	ug/L	4	Standard
	Tl	205	1176.7	6.0	**0.1515**	0.008	5.0	ug/L	2	Standard
	Pb	206	21429.5	1.2	**3.6210**	0.069	1.9	ug/L	240	Standard
	Pb	207	15728.2	0.6	**2.9798**	0.021	0.7	ug/L	198	Standard
	Pb	208	70387.8	1.2	**3.3811**	0.084	2.5	ug/L	883	Standard
	U	238	18087.1	1.2	**2.4141**	0.054	2.2	ug/L	3	Standard
>	**Bi**	209	569388.1	1.3				ug/L	523984	Standard

Sample ID: L1606138402
Report Date/Time: Wednesday, June 29, 2016 19:47:41
Page 1

Approved: July 04, 2016

Na	23	5.0	173.2	**43.9188**	76.061	173.2	mg/L	0	Standard
Mg	24	210.0	6.3	**0.1429**	0.013	9.0	mg/L	37	Standard
K	39	275.0	8.3	**0.3521**	0.037	10.4	mg/L	38	Standard
Ca	43	148.3	19.8	**4.6742**	1.471	31.5	mg/L	10	Standard
Fe	54	12478.7	2.5	**7.9817**	0.313	3.9	mg/L	179	Standard
Fe	57	3430.4	4.1	**7.8688**	0.221	2.8	mg/L	335	Standard
> **Sc-1**	45	79546.6	1.7				mg/L	61029	Standard
Cl	35	26111.8	0.7				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	2043.5	3.5				ug/L	1547	Standard
P	31	19195.5	2.0				ug/L	15286	Standard
S	34	2368.5	5.4				ug/L	1572	Standard
Sr	88	153.3	13.2				ug/L	73	Standard
C	12	70.0	14.3				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	140.0	18.9				mg/L	3	Standard
Dy	164	65719.4	1.8				mg/L	6	Standard
Ho-1	165	47698.3	1.5				mg/L	10	Standard
Er	166	45979.7	1.2				mg/L	13	Standard
I	127	5242.6	1.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		105.530	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.553	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138402
Report Date/Time: Wednesday, June 29, 2016 19:47:41
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	113.885
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.665
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138404
Sample Date/Time: Wednesday, June 29, 2016 19:48:35
Number of Replicates: 3
Autosampler Position: 405
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80955.8	1.9				ug/L	74699	Standard
	Be	9	880.0	13.6	**0.9302**	0.126	13.6	ug/L	8	Standard
	Al	27	4591606.5	1.6	**28.8172**	0.235	0.8	ug/L	1493	Standard
	Sc	45	81773.4	1.2				ug/L	61029	Standard
	Ti	47	8450.7	1.8	**19.2499**	0.245	1.3	ug/L	35	Standard
	V	51	185955.1	1.6	**21.1827**	0.308	1.5	ug/L	1435	Standard
	Cr	52	431888.4	0.7	**49.5778**	0.407	0.8	ug/L	7635	Standard
	Cr	53	54036.2	1.9	**50.4433**	1.377	2.7	ug/L	363	Standard
	Mn	55	1979621.7	0.4	**246.7487**	3.142	1.3	ug/L	747	Standard
	Co	59	89211.3	2.3	**11.6929**	0.262	2.2	ug/L	234	Standard
	Ni	60	241828.7	1.1	**112.0223**	0.265	0.2	ug/L	215	Standard
	Cu	65	79842.9	1.5	**37.1424**	0.248	0.7	ug/L	104	Standard
	Zn	66	363879.1	1.7	**288.7674**	1.874	0.6	ug/L	363	Standard
>	**Ge**	72	570278.1	1.3				ug/L	495142	Standard
	As	75	11266.0	1.4	**8.7923**	0.037	0.4	ug/L	-50	Standard
	Se	82	562.8	1.7	**4.3149**	0.122	2.8	ug/L	14	Standard
	Se-1	77	468.7	10.3	**4.9506**	0.587	11.9	ug/L	42	Standard
>	**Ga**	71	6996.6	2.8				mg/L	42	Standard
	Rb	85	75201.3	3.4				ug/L	17	Standard
	Y	89	1637404.7	1.2				ug/L	417098	Standard
>	**Rh**	103	40.0	12.5				ug/L	13	Standard
	Mo	98	6065.7	2.4	**2.0076**	0.060	3.0	ug/L	11	Standard
	Ag	107	921.4	1.6	**0.1122**	0.003	2.7	ug/L	68	Standard
	Cd	111	1933.3	0.9	**0.8101**	0.015	1.9	mg/L	2	Standard
	Cd	114	4671.7	0.8	**0.8037**	0.006	0.7	ug/L	12	Standard
>	**In**	115	664089.4	1.0				ug/L	587441	Standard
	Sn	118	708.3	7.1	**0.0125**	0.007	53.2	ug/L	497	Standard
	Sb	123	171.2	9.0	**0.0282**	0.003	10.1	ug/L	78	Standard
	Ba	135	217409.8	1.0	**93.9488**	0.914	1.0	ug/L	24	Standard
	Ce	140	744689.2	0.9				ug/L	25	Standard
>	**Tb**	159	1077211.2	1.7				ug/L	939701	Standard
	Ho	165	62662.7	1.1				ug/L	10	Standard
	Tl	203	2865.3	1.3	**0.3204**	0.004	1.2	ug/L	4	Standard
	Tl	205	2535.2	5.2	**0.3286**	0.016	4.9	ug/L	2	Standard
	Pb	206	40598.7	0.9	**6.9743**	0.127	1.8	ug/L	240	Standard
	Pb	207	29975.3	2.2	**5.7854**	0.186	3.2	ug/L	198	Standard
	Pb	208	132785.8	0.6	**6.4904**	0.091	1.4	ug/L	883	Standard
	U	238	27584.4	0.2	**3.7106**	0.059	1.6	ug/L	3	Standard
>	**Bi**	209	565317.9	1.5				ug/L	523984	Standard

Sample ID: L1606138404
Report Date/Time: Wednesday, June 29, 2016 19:50:52
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**14.1527**	24.505	173.1	mg/L	0	Standard
Mg	24	243.3	11.7	**0.1691**	0.025	14.5	mg/L	37	Standard
K	39	461.7	5.5	**0.6478**	0.049	7.5	mg/L	38	Standard
Ca	43	205.0	4.9	**7.1444**	0.559	7.8	mg/L	10	Standard
Fe	54	23854.6	1.3	**14.9557**	0.059	0.4	mg/L	179	Standard
Fe	57	6316.3	1.9	**14.9472**	0.123	0.8	mg/L	335	Standard
> **Sc-1**	45	81773.4	1.2				mg/L	61029	Standard
Cl	35	26395.0	1.0				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	8435.7	8.5				ug/L	1547	Standard
P	31	20819.3	1.9				ug/L	15286	Standard
S	34	2495.2	1.6				ug/L	1572	Standard
Sr	88	186.7	12.1				ug/L	73	Standard
C	12	26.7	21.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	520.0	8.4				mg/L	3	Standard
Dy	164	85733.4	1.4				mg/L	6	Standard
Ho-1	165	62662.7	1.1				mg/L	10	Standard
Er	166	61347.3	2.7				mg/L	13	Standard
I	127	12962.2	2.6				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		108.376	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		115.175	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606138404
Report Date/Time: Wednesday, June 29, 2016 19:50:52
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.048
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.888
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ni 60 Upper, S, EEE	Ni	60	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 19:51:49
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	79632.1	2.5				ug/L	74699	Standard
	Be	9	53037.7	1.9	**56.5448**	1.909	3.4	ug/L	8	Standard
	Al	27	8288547.2	1.5	**52.9043**	0.899	1.7	ug/L	1493	Standard
	Sc	45	68251.7	0.4				ug/L	61029	Standard
	Ti	47	44871.7	1.2	**102.9405**	2.024	2.0	ug/L	35	Standard
	V	51	450952.2	1.7	**51.7787**	1.350	2.6	ug/L	1435	Standard
	Cr	52	441674.6	0.8	**50.8503**	0.824	1.6	ug/L	7635	Standard
	Cr	53	54163.4	3.4	**50.6767**	1.716	3.4	ug/L	363	Standard
	Mn	55	408648.7	3.2	**50.9466**	2.025	4.0	ug/L	747	Standard
	Co	59	404996.9	2.1	**53.3495**	1.595	3.0	ug/L	234	Standard
	Ni	60	109209.9	1.2	**50.6487**	1.081	2.1	ug/L	215	Standard
	Cu	65	108068.7	1.0	**50.4372**	0.990	2.0	ug/L	104	Standard
	Zn	66	63914.2	1.2	**50.4711**	0.538	1.1	ug/L	363	Standard
>	Ge	72	568913.3	1.0				ug/L	495142	Standard
	As	75	64596.5	1.7	**50.2573**	1.198	2.4	ug/L	-50	Standard
	Se	82	6434.6	2.8	**50.8949**	1.678	3.3	ug/L	14	Standard
	Se-1	77	4404.6	0.9	**51.8699**	0.916	1.8	ug/L	42	Standard
>	Ga	71	81.7	19.7				mg/L	42	Standard
	Rb	85	868.4	8.5				ug/L	17	Standard
	Y	89	484608.1	0.7				ug/L	417098	Standard
>	Rh	103	25.0	40.0				ug/L	13	Standard
	Mo	98	300990.3	2.1	**101.3882**	1.976	1.9	ug/L	11	Standard
	Ag	107	359716.1	0.6	**49.5455**	0.289	0.6	ug/L	68	Standard
	Cd	111	115140.5	0.5	**49.4382**	0.205	0.4	mg/L	2	Standard
	Cd	114	277323.3	2.2	**48.5511**	1.074	2.2	ug/L	12	Standard
>	In	115	654244.8	0.1				ug/L	587441	Standard
	Sn	118	326410.3	0.9	**50.0768**	0.438	0.9	ug/L	497	Standard
	Sb	123	257734.5	0.6	**49.5626**	0.244	0.5	ug/L	78	Standard
	Ba	135	107937.1	0.2	**47.3350**	0.075	0.2	ug/L	24	Standard
	Ce	140	156.7	33.8				ug/L	25	Standard
>	Tb	159	993573.1	1.0				ug/L	939701	Standard
	Ho	165	18.3	41.7				ug/L	10	Standard
	Tl	203	442343.3	0.6	**51.3703**	0.497	1.0	ug/L	4	Standard
	Tl	205	397679.6	3.2	**53.2177**	1.929	3.6	ug/L	2	Standard
	Pb	206	282572.7	0.8	**50.5712**	0.781	1.5	ug/L	240	Standard
	Pb	207	252130.0	1.0	**50.7898**	0.969	1.9	ug/L	198	Standard
	Pb	208	1034503.8	1.3	**52.7468**	1.180	2.2	ug/L	883	Standard
	U	238	372072.1	1.5	**51.7486**	1.271	2.5	ug/L	3	Standard
>	Bi	209	547364.2	0.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:54:05
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.1101**	29.627	173.2	mg/L	0	Standard
Mg	24	4437.3	1.9	**5.0568**	0.105	2.1	mg/L	37	Standard
K	39	2166.8	4.9	**4.1677**	0.193	4.6	mg/L	38	Standard
Ca	43	105.0	40.7	**3.4097**	2.426	71.2	mg/L	10	Standard
Fe	54	6964.2	0.4	**5.1415**	0.037	0.7	mg/L	179	Standard
Fe	57	2053.5	5.5	**5.1659**	0.318	6.1	mg/L	335	Standard
> **Sc-1**	45	68251.7	0.4				mg/L	61029	Standard
Cl	35	27215.8	2.7				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	1920.1	6.9				ug/L	1547	Standard
P	31	20039.9	1.0				ug/L	15286	Standard
S	34	2546.9	7.4				ug/L	1572	Standard
Sr	88	71.7	29.0				ug/L	73	Standard
C	12	70.0	24.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	11.7	100.8				mg/L	6	Standard
Ho-1	165	18.3	41.7				mg/L	10	Standard
Er	166	33.3	34.6				mg/L	13	Standard
I	127	1696.8	22.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	113.090		
Al	27	105.809		
Sc	45			
Ti	47	102.941		
V	51	103.557		
Cr	52	101.701		
Cr	53			
Mn	55	101.893		
Co	59	106.699		
Ni	60	101.297		
Cu	65	100.874		
Zn	66	100.942		
> Ge	72		114.899	
As	75	100.515		
Se	82	101.790		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:54:05
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.388		
Ag	107	99.091		
Cd	111	98.876		
Cd	114			
In	115		111.372	
Sn	118	100.154		
Sb	123	99.125		
Ba	135	94.670		
Ce	140			
Tb	159			
Ho	165			
Tl	203	102.741		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.494		
U	238	103.497		
Bi	209		104.462	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:54:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 19:55:00
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78377.2	0.9				ug/L	74699	Standard
	Be	9	16.7	17.3	**0.0256**	0.003	11.6	ug/L	8	Standard
	Al	27	3955.5	8.5	**0.0123**	0.002	16.7	ug/L	1493	Standard
	Sc	45	70123.6	2.1				ug/L	61029	Standard
	Ti	47	44.3	19.2	**-0.0067**	0.022	331.3	ug/L	35	Standard
	V	51	1494.9	5.1	**-0.0231**	0.005	19.5	ug/L	1435	Standard
	Cr	52	7869.7	1.7	**-0.0976**	0.038	39.0	ug/L	7635	Standard
	Cr	53	438.3	4.0	**0.0839**	0.027	31.8	ug/L	363	Standard
	Mn	55	1323.4	11.3	**0.0175**	0.017	96.4	ug/L	747	Standard
	Co	59	292.0	15.1	**0.0012**	0.005	416.2	ug/L	234	Standard
	Ni	60	231.7	5.0	**-0.0160**	0.006	40.4	ug/L	215	Standard
	Cu	65	142.7	7.1	**-0.0369**	0.003	9.0	ug/L	104	Standard
	Zn	66	477.7	7.4	**-0.0736**	0.020	26.8	ug/L	363	Standard
>	**Ge**	72	568079.4	2.5				ug/L	495142	Standard
	As	75	-36.3	34.6	**0.0313**	0.010	31.8	ug/L	-50	Standard
	Se	82	17.1	26.2	**-0.0026**	0.033	1282.0	ug/L	14	Standard
	Se-1	77	54.3	9.3	**0.0268**	0.068	253.5	ug/L	42	Standard
>	**Ga**	71	40.0	12.5				mg/L	42	Standard
	Rb	85	38.3	19.9				ug/L	17	Standard
	Y	89	483139.1	0.9				ug/L	417098	Standard
>	**Rh**	103	11.7	99.0				ug/L	13	Standard
	Mo	98	291.7	36.0	**0.0908**	0.037	40.3	ug/L	11	Standard
	Ag	107	161.3	21.3	**0.0089**	0.005	57.1	ug/L	68	Standard
	Cd	111	21.8	28.2	**0.0012**	0.002	201.3	mg/L	2	Standard
	Cd	114	48.0	10.4	**0.0061**	0.001	15.7	ug/L	12	Standard
>	**In**	115	669206.2	1.8				ug/L	587441	Standard
	Sn	118	881.7	18.2	**0.0375**	0.022	58.1	ug/L	497	Standard
	Sb	123	447.3	19.7	**0.0797**	0.015	18.9	ug/L	78	Standard
	Ba	135	111.0	14.0	**0.0316**	0.007	21.6	ug/L	24	Standard
	Ce	140	185.0	42.5				ug/L	25	Standard
>	**Tb**	159	1027080.2	0.9				ug/L	939701	Standard
	Ho	165	28.3	10.2				ug/L	10	Standard
	Tl	203	102.0	27.3	**0.0095**	0.003	32.1	ug/L	4	Standard
	Tl	205	76.7	13.6	**0.0100**	0.001	13.1	ug/L	2	Standard
	Pb	206	326.3	10.0	**-0.0155**	0.005	35.3	ug/L	240	Standard
	Pb	207	292.3	4.6	**-0.0139**	0.002	17.4	ug/L	198	Standard
	Pb	208	1226.7	4.7	**-0.0145**	0.003	18.0	ug/L	883	Standard
	U	238	78.7	31.6	**0.0060**	0.003	54.4	ug/L	3	Standard
>	**Bi**	209	574845.6	0.5				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:57:17
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	34.4	**-0.0124**	0.019	155.4	mg/L	37	Standard
K	39	46.7	44.6	**-0.0238**	0.042	178.2	mg/L	38	Standard
Ca	43	26.7	57.3	**-1.0457**	0.868	83.0	mg/L	10	Standard
Fe	54	161.1	25.9	**-0.0190**	0.033	173.6	mg/L	179	Standard
Fe	57	251.7	19.3	**-0.3314**	0.129	38.8	mg/L	335	Standard
Sc-1	45	70123.6	2.1				mg/L	61029	Standard
Cl	35	27868.9	0.6				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	2040.1	1.8				ug/L	1547	Standard
P	31	20417.1	4.9				ug/L	15286	Standard
S	34	2425.2	3.5				ug/L	1572	Standard
Sr	88	45.0	40.1				ug/L	73	Standard
C	12	50.0	52.9				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	11.9	129.9				mg/L	6	Standard
Ho-1	165	28.3	10.2				mg/L	10	Standard
Er	166	30.0	57.7				mg/L	13	Standard
I	127	1460.1	42.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.731	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:57:17
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.919
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.707
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:57:17
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138406
Sample Date/Time: Wednesday, June 29, 2016 19:58:13
Number of Replicates: 3
Autosampler Position: 406
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77039.0	4.7				ug/L	74699	Standard
	Be	9	588.3	14.0	**0.6548**	0.074	11.3	ug/L	8	Standard
	Al	27	832772.7	0.7	**5.4883**	0.227	4.1	ug/L	1493	Standard
	Sc	45	75047.3	4.1				ug/L	61029	Standard
	Ti	47	6368.7	0.7	**14.6522**	0.311	2.1	ug/L	35	Standard
	V	51	122665.5	0.9	**14.0720**	0.287	2.0	ug/L	1435	Standard
	Cr	52	293324.5	0.5	**33.7422**	0.348	1.0	ug/L	7635	Standard
	Cr	53	35827.3	1.9	**33.7165**	0.271	0.8	ug/L	363	Standard
	Mn	55	1790039.2	2.5	**225.6481**	2.955	1.3	ug/L	747	Standard
	Co	59	59513.4	2.6	**7.8774**	0.097	1.2	ug/L	234	Standard
	Ni	60	162230.6	1.1	**75.9835**	0.734	1.0	ug/L	215	Standard
	Cu	65	58756.0	0.7	**27.6257**	0.286	1.0	ug/L	104	Standard
	Zn	66	213240.6	1.5	**171.0057**	1.221	0.7	ug/L	363	Standard
>	**Ge**	72	563753.3	1.5				ug/L	495142	Standard
	As	75	6354.9	1.3	**5.0439**	0.138	2.7	ug/L	-50	Standard
	Se	82	268.5	3.3	**2.0114**	0.102	5.1	ug/L	14	Standard
	Se-1	77	256.0	4.1	**2.4577**	0.168	6.8	ug/L	42	Standard
>	**Ga**	71	4010.5	2.8				mg/L	42	Standard
	Rb	85	39651.9	3.3				ug/L	17	Standard
	Y	89	1622063.0	2.1				ug/L	417098	Standard
>	**Rh**	103	33.3	8.7				ug/L	13	Standard
	Mo	98	4372.5	0.7	**1.4952**	0.034	2.3	ug/L	11	Standard
	Ag	107	608.3	2.2	**0.0725**	0.003	3.7	ug/L	68	Standard
	Cd	111	1369.4	3.4	**0.5916**	0.033	5.6	mg/L	2	Standard
	Cd	114	3220.4	7.1	**0.5731**	0.054	9.4	ug/L	12	Standard
>	**In**	115	642314.3	2.8				ug/L	587441	Standard
	Sn	118	866.7	3.3	**0.0411**	0.007	16.9	ug/L	497	Standard
	Sb	123	475.1	9.4	**0.0890**	0.010	11.7	ug/L	78	Standard
	Ba	135	210364.7	1.0	**94.0146**	1.722	1.8	ug/L	24	Standard
	Ce	140	620778.1	1.3				ug/L	25	Standard
>	**Tb**	159	1035282.0	2.1				ug/L	939701	Standard
	Ho	165	58499.4	0.8				ug/L	10	Standard
	Tl	203	1670.8	3.4	**0.1930**	0.007	3.5	ug/L	4	Standard
	Tl	205	1583.4	5.7	**0.2129**	0.013	6.3	ug/L	2	Standard
	Pb	206	23886.1	1.1	**4.2252**	0.039	0.9	ug/L	240	Standard
	Pb	207	17508.1	0.9	**3.4746**	0.041	1.2	ug/L	198	Standard
	Pb	208	79776.0	1.1	**4.0140**	0.064	1.6	ug/L	883	Standard
	U	238	22400.5	0.6	**3.1225**	0.026	0.8	ug/L	3	Standard
>	**Bi**	209	545342.1	0.8				ug/L	523984	Standard

Sample ID: L1606138406
Report Date/Time: Wednesday, June 29, 2016 20:00:30
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	190.0	5.3	**0.1344**	0.009	6.3	mg/L	37	Standard
K	39	241.7	12.5	**0.3203**	0.051	16.0	mg/L	38	Standard
Ca	43	168.3	3.4	**6.1267**	0.214	3.5	mg/L	10	Standard
Fe	54	12517.1	3.0	**8.4992**	0.364	4.3	mg/L	179	Standard
Fe	57	3428.7	4.4	**8.4118**	0.531	6.3	mg/L	335	Standard
> **Sc-1**	45	75047.3	4.1				mg/L	61029	Standard
Cl	35	26434.4	2.8				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2270.2	3.9				ug/L	1547	Standard
P	31	19440.8	0.5				ug/L	15286	Standard
S	34	2256.8	6.7				ug/L	1572	Standard
Sr	88	170.0	33.4				ug/L	73	Standard
C	12	76.7	54.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	300.0	18.6				mg/L	3	Standard
Dy	164	81135.2	0.1				mg/L	6	Standard
Ho-1	165	58499.4	0.8				mg/L	10	Standard
Er	166	56983.6	0.5				mg/L	13	Standard
I	127	5462.7	4.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.133	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.857	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138406
Report Date/Time: Wednesday, June 29, 2016 20:00:30
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	109.341
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.076
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138408
Sample Date/Time: Wednesday, June 29, 2016 20:01:24
Number of Replicates: 3
Autosampler Position: 407
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79667.3	1.9				ug/L	74699	Standard
	Be	9	540.0	10.4	**0.5836**	0.070	11.9	ug/L	8	Standard
	Al	27	892686.6	1.0	**5.6842**	0.164	2.9	ug/L	1493	Standard
	Sc	45	78111.0	3.7				ug/L	61029	Standard
	Ti	47	6171.3	0.9	**13.8487**	0.141	1.0	ug/L	35	Standard
	V	51	131186.8	0.8	**14.6946**	0.149	1.0	ug/L	1435	Standard
	Cr	52	291322.4	2.4	**32.6724**	0.743	2.3	ug/L	7635	Standard
	Cr	53	35120.7	4.3	**32.2467**	1.471	4.6	ug/L	363	Standard
	Mn	55	2233691.8	2.2	**274.8758**	6.127	2.2	ug/L	747	Standard
	Co	59	65245.2	1.7	**8.4325**	0.147	1.7	ug/L	234	Standard
	Ni	60	190260.1	2.2	**86.9887**	2.023	2.3	ug/L	215	Standard
	Cu	65	66483.6	1.3	**30.5171**	0.399	1.3	ug/L	104	Standard
	Zn	66	187493.0	1.2	**146.6893**	2.307	1.6	ug/L	363	Standard
>	**Ge**	72	577612.7	0.6				ug/L	495142	Standard
	As	75	6445.8	1.9	**4.9929**	0.116	2.3	ug/L	-50	Standard
	Se	82	222.3	2.5	**1.5982**	0.037	2.3	ug/L	14	Standard
	Se-1	77	243.0	1.1	**2.2298**	0.039	1.8	ug/L	42	Standard
>	**Ga**	71	4360.6	1.0				mg/L	42	Standard
	Rb	85	41170.9	1.0				ug/L	17	Standard
	Y	89	1636669.5	0.9				ug/L	417098	Standard
>	**Rh**	103	33.3	22.9				ug/L	13	Standard
	Mo	98	4750.5	1.5	**1.5922**	0.009	0.6	ug/L	11	Standard
	Ag	107	651.7	5.1	**0.0767**	0.003	4.5	ug/L	68	Standard
	Cd	111	1476.5	1.8	**0.6252**	0.009	1.4	mg/L	2	Standard
	Cd	114	3537.9	6.1	**0.6161**	0.029	4.7	ug/L	12	Standard
>	**In**	115	655238.7	1.5				ug/L	587441	Standard
	Sn	118	965.0	3.4	**0.0534**	0.004	7.7	ug/L	497	Standard
	Sb	123	268.7	11.3	**0.0474**	0.005	10.8	ug/L	78	Standard
	Ba	135	196413.8	1.6	**86.0176**	0.137	0.2	ug/L	24	Standard
	Ce	140	596487.3	1.9				ug/L	25	Standard
>	**Tb**	159	1073225.5	1.0				ug/L	939701	Standard
	Ho	165	61178.3	1.7				ug/L	10	Standard
	Tl	203	2456.5	1.8	**0.2718**	0.001	0.3	ug/L	4	Standard
	Tl	205	2150.2	4.8	**0.2763**	0.017	6.1	ug/L	2	Standard
	Pb	206	24200.3	1.9	**4.0880**	0.010	0.3	ug/L	240	Standard
	Pb	207	17656.3	0.7	**3.3464**	0.078	2.3	ug/L	198	Standard
	Pb	208	79187.0	3.1	**3.8025**	0.046	1.2	ug/L	883	Standard
	U	238	22940.4	1.4	**3.0557**	0.038	1.2	ug/L	3	Standard
>	**Bi**	209	570736.3	2.0				ug/L	523984	Standard

Sample ID: L1606138408
Report Date/Time: Wednesday, June 29, 2016 20:03:41
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**28.7877**	49.853	173.2	mg/L	0	Standard
Mg	24	161.7	15.3	**0.0975**	0.019	19.1	mg/L	37	Standard
K	39	273.3	12.8	**0.3583**	0.064	17.7	mg/L	38	Standard
Ca	43	191.7	9.9	**6.9329**	0.825	11.9	mg/L	10	Standard
Fe	54	12889.7	4.6	**8.4015**	0.298	3.5	mg/L	179	Standard
Fe	57	3420.4	5.7	**8.0292**	0.829	10.3	mg/L	335	Standard
Sc-1	45	78111.0	3.7				mg/L	61029	Standard
Cl	35	25108.8	1.1				ug/L	20195	Standard
Kr	83	1.0	173.2				ug/L	4	Standard
Br	81	1990.1	4.1				ug/L	1547	Standard
P	31	18960.2	2.1				ug/L	15286	Standard
S	34	2216.8	7.1				ug/L	1572	Standard
Sr	88	153.3	28.5				ug/L	73	Standard
C	12	110.0	9.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	313.3	12.9				mg/L	3	Standard
Dy	164	83986.8	2.5				mg/L	6	Standard
Ho-1	165	61178.3	1.7				mg/L	10	Standard
Er	166	57776.6	0.9				mg/L	13	Standard
I	127	4875.8	1.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		106.651	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		116.656	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138408
Report Date/Time: Wednesday, June 29, 2016 20:03:41
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	111.541
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.922
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606138408
Report Date/Time: Wednesday, June 29, 2016 20:03:41
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138410
Sample Date/Time: Wednesday, June 29, 2016 20:04:36
Number of Replicates: 3
Autosampler Position: 408
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82222.6	2.1				ug/L	74699	Standard
	Be	9	466.7	0.6	**0.4893**	0.010	2.0	ug/L	8	Standard
	Al	27	798979.6	1.9	**4.9264**	0.080	1.6	ug/L	1493	Standard
	Sc	45	79640.5	1.7				ug/L	61029	Standard
	Ti	47	5792.4	2.5	**13.1770**	0.593	4.5	ug/L	35	Standard
	V	51	95139.7	1.4	**10.7518**	0.246	2.3	ug/L	1435	Standard
	Cr	52	279542.5	2.0	**31.7533**	0.664	2.1	ug/L	7635	Standard
	Cr	53	34038.2	1.3	**31.6794**	0.880	2.8	ug/L	363	Standard
	Mn	55	1053790.0	0.3	**131.3975**	3.080	2.3	ug/L	747	Standard
	Co	59	35027.1	0.6	**4.5723**	0.080	1.8	ug/L	234	Standard
	Ni	60	156240.8	1.2	**72.3985**	1.638	2.3	ug/L	215	Standard
	Cu	65	48550.0	0.3	**22.5657**	0.405	1.8	ug/L	104	Standard
	Zn	66	257712.2	0.3	**204.5897**	4.503	2.2	ug/L	363	Standard
>	**Ge**	72	569889.2	2.0				ug/L	495142	Standard
	As	75	6132.6	2.6	**4.8186**	0.190	4.0	ug/L	-50	Standard
	Se	82	361.1	9.2	**2.7244**	0.314	11.5	ug/L	14	Standard
	Se-1	77	331.7	7.4	**3.3212**	0.233	7.0	ug/L	42	Standard
>	**Ga**	71	4152.2	2.1				mg/L	42	Standard
	Rb	85	52623.0	3.5				ug/L	17	Standard
	Y	89	1425089.1	0.3				ug/L	417098	Standard
>	**Rh**	103	33.3	8.7				ug/L	13	Standard
	Mo	98	3927.9	0.3	**1.2648**	0.018	1.4	ug/L	11	Standard
	Ag	107	467.0	2.9	**0.0489**	0.003	5.2	ug/L	68	Standard
	Cd	111	1169.0	3.4	**0.4740**	0.011	2.3	mg/L	2	Standard
	Cd	114	2891.7	1.0	**0.4840**	0.007	1.5	ug/L	12	Standard
>	**In**	115	681463.1	1.2				ug/L	587441	Standard
	Sn	118	755.0	5.0	**0.0168**	0.007	41.1	ug/L	497	Standard
	Sb	123	219.3	13.0	**0.0363**	0.005	13.2	ug/L	78	Standard
	Ba	135	73803.3	1.7	**31.0670**	0.270	0.9	ug/L	24	Standard
	Ce	140	554709.1	1.1				ug/L	25	Standard
>	**Tb**	159	1094015.2	1.4				ug/L	939701	Standard
	Ho	165	50373.6	1.6				ug/L	10	Standard
	Tl	203	1120.7	1.3	**0.1214**	0.002	2.0	ug/L	4	Standard
	Tl	205	941.7	7.2	**0.1194**	0.008	6.7	ug/L	2	Standard
	Pb	206	24028.0	1.2	**4.0020**	0.045	1.1	ug/L	240	Standard
	Pb	207	17504.4	0.3	**3.2701**	0.054	1.6	ug/L	198	Standard
	Pb	208	78967.6	0.8	**3.7402**	0.091	2.4	ug/L	883	Standard
	U	238	25248.7	0.7	**3.3178**	0.077	2.3	ug/L	3	Standard
>	**Bi**	209	578682.9	1.6				ug/L	523984	Standard

Sample ID: L1606138410
Report Date/Time: Wednesday, June 29, 2016 20:06:53
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	188.3	33.0	**0.1211**	0.061	50.1	mg/L	37	Standard
K	39	248.3	3.1	**0.3067**	0.020	6.4	mg/L	38	Standard
Ca	43	173.3	7.3	**5.8747**	0.743	12.6	mg/L	10	Standard
Fe	54	14056.9	1.9	**8.9942**	0.034	0.4	mg/L	179	Standard
Fe	57	3822.2	6.9	**8.8892**	0.835	9.4	mg/L	335	Standard
> **Sc-1**	45	79640.5	1.7				mg/L	61029	Standard
Cl	35	25424.6	1.4				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	2053.5	6.1				ug/L	1547	Standard
P	31	20016.5	1.3				ug/L	15286	Standard
S	34	2301.8	5.1				ug/L	1572	Standard
Sr	88	141.7	32.0				ug/L	73	Standard
C	12	56.7	36.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	133.3	11.5				mg/L	3	Standard
Dy	164	72545.2	2.7				mg/L	6	Standard
Ho-1	165	50373.6	1.6				mg/L	10	Standard
Er	166	47160.0	0.9				mg/L	13	Standard
I	127	6799.9	2.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.072	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.096	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138410
Report Date/Time: Wednesday, June 29, 2016 20:06:53
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	116.005
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	110.439
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138412
Sample Date/Time: Wednesday, June 29, 2016 20:07:47
Number of Replicates: 3
Autosampler Position: 409
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81391.5	3.4				ug/L	74699	Standard
	Be	9	565.0	6.9	**0.5962**	0.020	3.4	ug/L	8	Standard
	Al	27	802320.4	2.6	**4.9984**	0.082	1.6	ug/L	1493	Standard
	Sc	45	77871.4	3.1				ug/L	61029	Standard
	Ti	47	8431.3	2.1	**19.3525**	0.252	1.3	ug/L	35	Standard
	V	51	136034.9	0.6	**15.5641**	0.178	1.1	ug/L	1435	Standard
	Cr	52	292890.6	0.7	**33.5584**	0.660	2.0	ug/L	7635	Standard
	Cr	53	35682.0	1.9	**33.4491**	0.599	1.8	ug/L	363	Standard
	Mn	55	2260782.0	0.5	**283.9966**	5.862	2.1	ug/L	747	Standard
	Co	59	82153.8	0.1	**10.8487**	0.184	1.7	ug/L	234	Standard
	Ni	60	172794.1	1.0	**80.6313**	1.536	1.9	ug/L	215	Standard
	Cu	65	63565.7	0.4	**29.7801**	0.432	1.5	ug/L	104	Standard
	Zn	66	251019.4	0.8	**200.6145**	2.591	1.3	ug/L	363	Standard
>	**Ge**	72	565970.9	1.7				ug/L	495142	Standard
	As	75	7636.9	2.3	**6.0250**	0.142	2.4	ug/L	-50	Standard
	Se	82	426.3	0.8	**3.2611**	0.084	2.6	ug/L	14	Standard
	Se-1	77	381.7	1.3	**3.9506**	0.115	2.9	ug/L	42	Standard
>	**Ga**	71	4992.5	2.6				mg/L	42	Standard
	Rb	85	46121.8	0.5				ug/L	17	Standard
	Y	89	1754842.6	1.7				ug/L	417098	Standard
>	**Rh**	103	43.3	35.3				ug/L	13	Standard
	Mo	98	5382.6	1.1	**1.7465**	0.044	2.5	ug/L	11	Standard
	Ag	107	637.3	4.9	**0.0719**	0.004	5.2	ug/L	68	Standard
	Cd	111	1430.5	6.2	**0.5853**	0.028	4.9	mg/L	2	Standard
	Cd	114	3407.8	7.3	**0.5746**	0.047	8.2	ug/L	12	Standard
>	**In**	115	677210.7	1.5				ug/L	587441	Standard
	Sn	118	780.0	4.5	**0.0212**	0.006	30.1	ug/L	497	Standard
	Sb	123	243.4	9.9	**0.0411**	0.005	12.6	ug/L	78	Standard
	Ba	135	199262.8	0.6	**84.4450**	1.189	1.4	ug/L	24	Standard
	Ce	140	678125.1	0.7				ug/L	25	Standard
>	**Tb**	159	1071624.9	1.1				ug/L	939701	Standard
	Ho	165	67051.4	0.5				ug/L	10	Standard
	Tl	203	1859.8	4.4	**0.2096**	0.007	3.5	ug/L	4	Standard
	Tl	205	1828.4	5.2	**0.2396**	0.011	4.7	ug/L	2	Standard
	Pb	206	33182.6	0.6	**5.7494**	0.057	1.0	ug/L	240	Standard
	Pb	207	24419.7	0.3	**4.7511**	0.035	0.7	ug/L	198	Standard
	Pb	208	111320.6	0.3	**5.4889**	0.050	0.9	ug/L	883	Standard
	U	238	25349.9	1.3	**3.4462**	0.014	0.4	ug/L	3	Standard
>	**Bi**	209	559224.1	1.0				ug/L	523984	Standard

Sample ID: L1606138412
Report Date/Time: Wednesday, June 29, 2016 20:10:04
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**30.2574**	26.235	86.7	mg/L	0	Standard
Mg	24	168.3	9.1	**0.1052**	0.014	12.9	mg/L	37	Standard
K	39	280.0	6.4	**0.3709**	0.027	7.3	mg/L	38	Standard
Ca	43	176.7	27.2	**6.2736**	2.632	42.0	mg/L	10	Standard
Fe	54	15625.1	2.7	**10.2448**	0.046	0.4	mg/L	179	Standard
Fe	57	4357.3	3.7	**10.5343**	0.341	3.2	mg/L	335	Standard
> **Sc-1**	45	77871.4	3.1				mg/L	61029	Standard
Cl	35	26050.4	2.2				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2310.2	5.7				ug/L	1547	Standard
P	31	19966.5	2.1				ug/L	15286	Standard
S	34	2301.8	1.2				ug/L	1572	Standard
Sr	88	181.7	8.4				ug/L	73	Standard
C	12	60.0	66.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	466.7	15.8				mg/L	3	Standard
Dy	164	94635.2	0.7				mg/L	6	Standard
Ho-1	165	67051.4	0.5				mg/L	10	Standard
Er	166	62943.8	1.5				mg/L	13	Standard
I	127	6689.8	6.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		108.960	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.305	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138412
Report Date/Time: Wednesday, June 29, 2016 20:10:04
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	115.281
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	106.725
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606138412
Report Date/Time: Wednesday, June 29, 2016 20:10:04
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 20:11:01
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78720.7	2.9				ug/L	74699	Standard
	Be	9	53925.8	1.9	**58.1477**	1.206	2.1	ug/L	8	Standard
	Al	27	8443805.5	1.2	**54.5410**	1.847	3.4	ug/L	1493	Standard
	Sc	45	68074.4	3.5				ug/L	61029	Standard
	Ti	47	45146.2	1.1	**105.2966**	1.372	1.3	ug/L	35	Standard
	V	51	452407.6	0.9	**52.8112**	0.533	1.0	ug/L	1435	Standard
	Cr	52	439865.6	0.2	**51.4989**	0.494	1.0	ug/L	7635	Standard
	Cr	53	55131.8	1.5	**52.4610**	1.191	2.3	ug/L	363	Standard
	Mn	55	411766.6	1.7	**52.1886**	1.083	2.1	ug/L	747	Standard
	Co	59	411137.6	1.8	**55.0576**	1.090	2.0	ug/L	234	Standard
	Ni	60	108563.6	1.0	**51.1869**	0.633	1.2	ug/L	215	Standard
	Cu	65	108514.2	0.8	**51.4879**	0.273	0.5	ug/L	104	Standard
	Zn	66	63813.1	1.5	**51.2390**	0.656	1.3	ug/L	363	Standard
>	**Ge**	72	559563.9	0.9				ug/L	495142	Standard
	As	75	65025.3	0.7	**51.4299**	0.351	0.7	ug/L	-50	Standard
	Se	82	6607.2	0.8	**53.1327**	0.267	0.5	ug/L	14	Standard
	Se-1	77	4509.3	0.5	**54.0117**	0.383	0.7	ug/L	42	Standard
>	**Ga**	71	68.3	11.2				mg/L	42	Standard
	Rb	85	971.7	11.9				ug/L	17	Standard
	Y	89	484276.8	2.3				ug/L	417098	Standard
>	**Rh**	103	33.3	52.7				ug/L	13	Standard
	Mo	98	305352.5	1.7	**102.8803**	2.601	2.5	ug/L	11	Standard
	Ag	107	366186.0	1.2	**50.4411**	0.533	1.1	ug/L	68	Standard
	Cd	111	118185.8	1.5	**50.7502**	0.681	1.3	mg/L	2	Standard
	Cd	114	288421.2	1.0	**50.5001**	0.711	1.4	ug/L	12	Standard
>	**In**	115	654225.0	1.4				ug/L	587441	Standard
	Sn	118	325668.1	1.2	**49.9654**	0.174	0.3	ug/L	497	Standard
	Sb	123	260726.4	2.1	**50.1401**	0.797	1.6	ug/L	78	Standard
	Ba	135	109583.9	1.3	**48.0616**	0.591	1.2	ug/L	24	Standard
	Ce	140	191.7	20.9				ug/L	25	Standard
>	**Tb**	159	983539.2	1.4				ug/L	939701	Standard
	Ho	165	21.7	13.3				ug/L	10	Standard
	Tl	203	449254.6	0.7	**52.6820**	0.493	0.9	ug/L	4	Standard
	Tl	205	410287.0	1.8	**55.4353**	1.032	1.9	ug/L	2	Standard
	Pb	206	287176.7	1.5	**51.8952**	0.724	1.4	ug/L	240	Standard
	Pb	207	254226.6	1.1	**51.7082**	0.492	1.0	ug/L	198	Standard
	Pb	208	1062593.6	0.7	**54.7051**	0.438	0.8	ug/L	883	Standard
	U	238	383746.1	0.8	**53.8863**	0.314	0.6	ug/L	3	Standard
>	**Bi**	209	542059.3	0.3				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:13:18
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.9421**	31.068	173.2	mg/L	0	Standard
Mg	24	4572.4	3.8	**5.2291**	0.219	4.2	mg/L	37	Standard
K	39	2235.2	5.1	**4.3138**	0.121	2.8	mg/L	38	Standard
Ca	43	108.3	35.8	**3.6601**	2.380	65.0	mg/L	10	Standard
Fe	54	7126.2	2.9	**5.2851**	0.313	5.9	mg/L	179	Standard
Fe	57	2081.8	4.0	**5.2741**	0.321	6.1	mg/L	335	Standard
Sc-1	45	68074.4	3.5				mg/L	61029	Standard
Cl	35	27726.7	0.9				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	1910.1	6.0				ug/L	1547	Standard
P	31	20502.2	2.0				ug/L	15286	Standard
S	34	2426.9	4.0				ug/L	1572	Standard
Sr	88	50.0	26.5				ug/L	73	Standard
C	12	73.3	20.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	15.9	37.3				mg/L	6	Standard
Ho-1	165	21.7	13.3				mg/L	10	Standard
Er	166	16.7	34.6				mg/L	13	Standard
I	127	1525.1	31.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	116.295		
Al	27	109.082		
Sc	45			
Ti	47	105.297		
V	51	105.622		
Cr	52	102.998		
Cr	53			
Mn	55	104.377		
Co	59	110.115		
Ni	60	102.374		
Cu	65	102.976		
Zn	66	102.478		
Ge	72		113.011	
As	75	102.860		
Se	82	106.265		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:13:18
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.880		
Ag	107	100.882		
Cd	111	101.500		
Cd	114			
In	115		111.369	
Sn	118	99.931		
Sb	123	100.280		
Ba	135	96.123		
Ce	140			
Tb	159			
Ho	165			
Tl	203	105.364		
Tl	205			
Pb	206			
Pb	207			
Pb	208	109.410		
U	238	107.773		
Bi	209		103.450	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:13:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 20:14:12
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79194.8	0.3				ug/L	74699	Standard
	Be	9	31.7	9.1	**0.0415**	0.003	7.4	ug/L	8	Standard
	Al	27	5331.0	39.7	**0.0208**	0.013	64.7	ug/L	1493	Standard
	Sc	45	70443.5	2.7				ug/L	61029	Standard
	Ti	47	45.0	15.6	**-0.0077**	0.016	206.1	ug/L	35	Standard
	V	51	1613.4	5.4	**-0.0133**	0.013	95.3	ug/L	1435	Standard
	Cr	52	7981.4	3.6	**-0.1047**	0.047	45.2	ug/L	7635	Standard
	Cr	53	418.3	25.3	**0.0570**	0.103	180.4	ug/L	363	Standard
	Mn	55	1741.1	30.3	**0.0658**	0.068	103.0	ug/L	747	Standard
	Co	59	378.7	17.0	**0.0117**	0.009	74.2	ug/L	234	Standard
	Ni	60	306.7	29.3	**0.0162**	0.043	265.9	ug/L	215	Standard
	Cu	65	193.0	22.7	**-0.0150**	0.021	143.0	ug/L	104	Standard
	Zn	66	576.0	19.1	**-0.0039**	0.093	2387.7	ug/L	363	Standard
>	**Ge**	72	580508.2	1.5				ug/L	495142	Standard
	As	75	-17.3	276.0	**0.0468**	0.036	77.6	ug/L	-50	Standard
	Se	82	19.7	47.6	**0.0162**	0.076	468.0	ug/L	14	Standard
	Se-1	77	62.3	8.8	**0.1062**	0.073	68.4	ug/L	42	Standard
>	**Ga**	71	43.3	24.0				mg/L	42	Standard
	Rb	85	53.3	27.1				ug/L	17	Standard
	Y	89	466004.6	0.7				ug/L	417098	Standard
>	**Rh**	103	18.3	41.7				ug/L	13	Standard
	Mo	98	326.5	17.1	**0.1045**	0.019	18.6	ug/L	11	Standard
	Ag	107	195.7	20.5	**0.0141**	0.005	38.2	ug/L	68	Standard
	Cd	111	49.8	49.3	**0.0134**	0.010	76.8	mg/L	2	Standard
	Cd	114	97.6	30.0	**0.0150**	0.005	33.6	ug/L	12	Standard
>	**In**	115	653349.0	0.9				ug/L	587441	Standard
	Sn	118	958.4	16.5	**0.0527**	0.023	43.7	ug/L	497	Standard
	Sb	123	468.2	24.3	**0.0858**	0.021	24.7	ug/L	78	Standard
	Ba	135	169.0	53.9	**0.0580**	0.039	67.5	ug/L	24	Standard
	Ce	140	383.3	72.7				ug/L	25	Standard
>	**Tb**	159	1008003.3	2.5				ug/L	939701	Standard
	Ho	165	33.3	56.8				ug/L	10	Standard
	Tl	203	138.0	27.1	**0.0136**	0.004	29.9	ug/L	4	Standard
	Tl	205	128.3	47.6	**0.0167**	0.008	47.6	ug/L	2	Standard
	Pb	206	364.7	12.5	**-0.0084**	0.008	93.5	ug/L	240	Standard
	Pb	207	321.7	11.0	**-0.0076**	0.006	84.7	ug/L	198	Standard
	Pb	208	1298.7	10.5	**-0.0103**	0.007	67.0	ug/L	883	Standard
	U	238	121.3	35.3	**0.0118**	0.006	48.9	ug/L	3	Standard
>	**Bi**	209	569222.6	0.7				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:16:29
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	31.7	50.8	**-0.0295**	0.019	63.3	mg/L	37	Standard
K	39	58.3	13.1	**-0.0021**	0.017	817.3	mg/L	38	Standard
Ca	43	26.7	28.6	**-1.0586**	0.442	41.7	mg/L	10	Standard
Fe	54	154.2	11.8	**-0.0252**	0.011	43.9	mg/L	179	Standard
Fe	57	258.3	4.9	**-0.3125**	0.056	18.0	mg/L	335	Standard
> **Sc-1**	45	70443.5	2.7				mg/L	61029	Standard
Cl	35	27476.9	1.3				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	1946.8	3.9				ug/L	1547	Standard
P	31	20592.3	4.9				ug/L	15286	Standard
S	34	2348.5	2.8				ug/L	1572	Standard
Sr	88	75.0	46.7				ug/L	73	Standard
C	12	43.3	13.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	74.1	39.2				mg/L	6	Standard
Ho-1	165	33.3	56.8				mg/L	10	Standard
Er	166	53.3	71.0				mg/L	13	Standard
I	127	1293.4	30.4				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		117.241	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:16:29
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.219
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.633
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:16:29
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBW 43 WG573291-03
Sample Date/Time: Wednesday, June 29, 2016 20:17:26
Number of Replicates: 3
Autosampler Position: 410
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78266.6	1.0				ug/L	74699	Standard
	Be	9	21.7	58.1	**0.0312**	0.014	44.4	ug/L	8	Standard
	Al	27	32901.2	14.4	**0.2003**	0.031	15.3	ug/L	1493	Standard
	Sc	45	68990.1	1.1				ug/L	61029	Standard
	Ti	47	59.7	25.6	**0.0271**	0.036	131.8	ug/L	35	Standard
	V	51	1740.6	13.9	**0.0036**	0.029	816.2	ug/L	1435	Standard
	Cr	52	10743.1	2.3	**0.2282**	0.033	14.5	ug/L	7635	Standard
	Cr	53	856.7	10.0	**0.4710**	0.086	18.2	ug/L	363	Standard
	Mn	55	4965.9	44.9	**0.4694**	0.280	59.7	ug/L	747	Standard
	Co	59	438.7	27.8	**0.0202**	0.016	81.0	ug/L	234	Standard
	Ni	60	464.7	17.9	**0.0906**	0.040	44.1	ug/L	215	Standard
	Cu	65	288.0	10.4	**0.0300**	0.015	49.7	ug/L	104	Standard
	Zn	66	2615.6	1.4	**1.6135**	0.041	2.5	ug/L	363	Standard
>	**Ge**	72	573366.1	0.8				ug/L	495142	Standard
	As	75	-37.9	81.3	**0.0303**	0.024	78.8	ug/L	-50	Standard
	Se	82	11.3	32.8	**-0.0488**	0.029	60.5	ug/L	14	Standard
	Se-1	77	61.0	9.1	**0.0990**	0.067	67.3	ug/L	42	Standard
>	**Ga**	71	31.7	9.1				mg/L	42	Standard
	Rb	85	100.0	26.5				ug/L	17	Standard
	Y	89	479627.1	2.2				ug/L	417098	Standard
>	**Rh**	103	1.7	173.2				ug/L	13	Standard
	Mo	98	155.9	47.8	**0.0477**	0.025	52.9	ug/L	11	Standard
	Ag	107	191.7	76.1	**0.0140**	0.020	145.7	ug/L	68	Standard
	Cd	111	21.5	58.8	**0.0014**	0.005	377.2	mg/L	2	Standard
	Cd	114	53.4	80.4	**0.0074**	0.008	102.3	ug/L	12	Standard
>	**In**	115	641242.2	1.6				ug/L	587441	Standard
	Sn	118	930.0	4.7	**0.0512**	0.008	15.1	ug/L	497	Standard
	Sb	123	286.5	19.6	**0.0520**	0.011	21.3	ug/L	78	Standard
	Ba	135	402.7	93.5	**0.1636**	0.167	102.4	ug/L	24	Standard
	Ce	140	951.7	120.0				ug/L	25	Standard
>	**Tb**	159	992014.1	1.0				ug/L	939701	Standard
	Ho	165	68.3	105.1				ug/L	10	Standard
	Tl	203	138.0	108.8	**0.0140**	0.017	122.9	ug/L	4	Standard
	Tl	205	136.7	67.7	**0.0182**	0.012	67.1	ug/L	2	Standard
	Pb	206	481.7	43.4	**0.0137**	0.037	272.5	ug/L	240	Standard
	Pb	207	395.0	43.8	**0.0084**	0.035	416.2	ug/L	198	Standard
	Pb	208	1692.0	26.2	**0.0108**	0.023	209.9	ug/L	883	Standard
	U	238	120.3	76.5	**0.0121**	0.013	104.7	ug/L	3	Standard
>	**Bi**	209	557465.1	0.7				ug/L	523984	Standard

Sample ID: PBW 43 WG573291-03
Report Date/Time: Wednesday, June 29, 2016 20:19:42
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	43.3	**-0.0120**	0.023	188.7	mg/L	37	Standard
K	39	41.7	73.3	**-0.0331**	0.059	178.0	mg/L	38	Standard
Ca	43	16.7	45.8	**-1.5897**	0.437	27.5	mg/L	10	Standard
Fe	54	190.8	15.9	**0.0047**	0.022	466.7	mg/L	179	Standard
Fe	57	281.7	18.5	**-0.2277**	0.154	67.5	mg/L	335	Standard
Sc-1	45	68990.1	1.1				mg/L	61029	Standard
Cl	35	27241.8	2.0				ug/L	20195	Standard
Kr	83	3.3	34.6				ug/L	4	Standard
Br	81	1960.1	2.7				ug/L	1547	Standard
P	31	19042.0	2.0				ug/L	15286	Standard
S	34	2271.8	2.5				ug/L	1572	Standard
Sr	88	70.0	14.3				ug/L	73	Standard
C	12	93.3	12.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	57.4	72.5				mg/L	6	Standard
Ho-1	165	68.3	105.1				mg/L	10	Standard
Er	166	53.3	75.8				mg/L	13	Standard
I	127	1241.7	37.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		104.776	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.798	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: PBW 43 WG573291-03
Report Date/Time: Wednesday, June 29, 2016 20:19:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.159
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.390
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 43 WG573291-03
Report Date/Time: Wednesday, June 29, 2016 20:19:42
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSW 43 WG573291-04
Sample Date/Time: Wednesday, June 29, 2016 20:20:37
Number of Replicates: 3
Autosampler Position: 411
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79286.9	1.2				ug/L	74699	Standard
	Be	9	55518.2	1.7	**59.4178**	0.565	1.0	ug/L	8	Standard
	Al	27	18234.3	3.1	**0.1035**	0.005	4.4	ug/L	1493	Standard
	Sc	45	71925.5	3.7				ug/L	61029	Standard
	Ti	47	66.7	12.9	**0.0416**	0.023	54.2	ug/L	35	Standard
	V	51	464385.2	1.4	**52.3659**	1.151	2.2	ug/L	1435	Standard
	Cr	52	468709.1	0.6	**53.0500**	1.578	3.0	ug/L	7635	Standard
	Cr	53	57400.2	0.6	**52.7689**	1.471	2.8	ug/L	363	Standard
	Mn	55	425738.4	0.7	**52.1249**	1.035	2.0	ug/L	747	Standard
	Co	59	411335.9	1.2	**53.2163**	1.483	2.8	ug/L	234	Standard
	Ni	60	114540.7	1.0	**52.1842**	1.705	3.3	ug/L	215	Standard
	Cu	65	115424.0	0.4	**52.9151**	1.312	2.5	ug/L	104	Standard
	Zn	66	68318.0	0.2	**53.0163**	1.377	2.6	ug/L	363	Standard
>	**Ge**	72	579422.8	2.7				ug/L	495142	Standard
	As	75	63888.1	1.5	**48.8227**	1.412	2.9	ug/L	-50	Standard
	Se	82	6548.4	1.5	**50.8872**	2.152	4.2	ug/L	14	Standard
	Se-1	77	4570.0	2.1	**52.8570**	0.847	1.6	ug/L	42	Standard
>	**Ga**	71	68.3	27.7				mg/L	42	Standard
	Rb	85	96.7	7.9				ug/L	17	Standard
	Y	89	471340.1	1.5				ug/L	417098	Standard
>	**Rh**	103	33.3	52.7				ug/L	13	Standard
	Mo	98	1844.9	1.6	**0.6204**	0.021	3.4	ug/L	11	Standard
	Ag	107	386699.9	0.4	**53.6468**	1.108	2.1	ug/L	68	Standard
	Cd	111	120754.4	0.7	**52.2256**	1.320	2.5	mg/L	2	Standard
	Cd	114	290744.9	1.2	**51.2713**	1.454	2.8	ug/L	12	Standard
>	**In**	115	649724.4	1.8				ug/L	587441	Standard
	Sn	118	890.0	5.4	**0.0430**	0.007	15.1	ug/L	497	Standard
	Sb	123	259622.2	1.1	**50.2896**	1.419	2.8	ug/L	78	Standard
	Ba	135	113320.2	1.2	**50.0586**	1.415	2.8	ug/L	24	Standard
	Ce	140	216.7	24.0				ug/L	25	Standard
>	**Tb**	159	1024515.0	4.0				ug/L	939701	Standard
	Ho	165	13.3	57.3				ug/L	10	Standard
	Tl	203	467714.3	0.2	**50.8953**	1.428	2.8	ug/L	4	Standard
	Tl	205	419104.4	1.0	**52.5470**	1.541	2.9	ug/L	2	Standard
	Pb	206	302725.5	0.9	**50.7682**	1.716	3.4	ug/L	240	Standard
	Pb	207	257785.2	0.8	**48.6571**	1.672	3.4	ug/L	198	Standard
	Pb	208	1092150.5	0.9	**52.1642**	0.990	1.9	ug/L	883	Standard
	U	238	389261.3	0.3	**50.7252**	1.507	3.0	ug/L	3	Standard
>	**Bi**	209	584427.6	2.7				ug/L	523984	Standard

Sample ID: LCSW 43 WG573291-04
Report Date/Time: Wednesday, June 29, 2016 20:22:54
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	56.7	20.4	**-0.0027**	0.015	535.8	mg/L	37	Standard
K	39	35.0	42.9	**-0.0491**	0.026	52.4	mg/L	38	Standard
Ca	43	18.3	15.7	**-1.5405**	0.154	10.0	mg/L	10	Standard
Fe	54	173.6	39.1	**-0.0145**	0.044	301.9	mg/L	179	Standard
Fe	57	231.7	7.6	**-0.4064**	0.037	9.2	mg/L	335	Standard
Sc-1	45	71925.5	3.7				mg/L	61029	Standard
Cl	35	25598.9	2.5				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1866.8	9.1				ug/L	1547	Standard
P	31	19263.9	1.4				ug/L	15286	Standard
S	34	2041.8	5.6				ug/L	1572	Standard
Sr	88	55.0	27.3				ug/L	73	Standard
C	12	76.7	37.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	19.2	93.9				mg/L	6	Standard
Ho-1	165	13.3	57.3				mg/L	10	Standard
Er	166	16.7	124.9				mg/L	13	Standard
I	127	1405.1	22.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		106.142	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		117.021	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 43 WG573291-04
Report Date/Time: Wednesday, June 29, 2016 20:22:54
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.602
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.535
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401 WG573291-01
Sample Date/Time: Wednesday, June 29, 2016 20:23:48
Number of Replicates: 3
Autosampler Position: 412
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81548.9	1.6				ug/L	74699	Standard
	Be	9	36.7	15.7	**0.0457**	0.006	12.1	ug/L	8	Standard
	Al	27	11596363.1	1.3	**72.2698**	0.564	0.8	ug/L	1493	Standard
	Sc	45	73884.7	1.3				ug/L	61029	Standard
	Ti	47	344.7	7.0	**0.7127**	0.059	8.2	ug/L	35	Standard
	V	51	2330.4	3.5	**0.0834**	0.015	17.8	ug/L	1435	Standard
	Cr	52	13306.8	2.8	**0.5995**	0.076	12.7	ug/L	7635	Standard
	Cr	53	1323.4	7.3	**0.9647**	0.110	11.4	ug/L	363	Standard
	Mn	55	1766896.9	0.7	**229.1909**	3.866	1.7	ug/L	747	Standard
	Co	59	11800.3	0.6	**1.5775**	0.021	1.4	ug/L	234	Standard
	Ni	60	3337.7	1.4	**1.4870**	0.026	1.7	ug/L	215	Standard
	Cu	65	729.0	2.5	**0.2502**	0.002	1.0	ug/L	104	Standard
	Zn	66	5843.1	1.1	**4.3791**	0.046	1.0	ug/L	363	Standard
>	**Ge**	72	548022.2	2.0				ug/L	495142	Standard
	As	75	368.2	8.1	**0.3569**	0.029	8.2	ug/L	-50	Standard
	Se	82	62.0	10.3	**0.3726**	0.060	16.1	ug/L	14	Standard
	Se-1	77	77.3	2.7	**0.3344**	0.023	6.7	ug/L	42	Standard
>	**Ga**	71	126.7	17.8				mg/L	42	Standard
	Rb	85	31524.4	1.4				ug/L	17	Standard
	Y	89	492847.4	3.2				ug/L	417098	Standard
>	**Rh**	103	60.0	14.4				ug/L	13	Standard
	Mo	98	169.7	8.5	**0.0488**	0.006	11.4	ug/L	11	Standard
	Ag	107	132.7	27.4	**0.0046**	0.005	112.7	ug/L	68	Standard
	Cd	111	39.0	30.5	**0.0081**	0.005	64.3	mg/L	2	Standard
	Cd	114	83.8	14.7	**0.0119**	0.002	19.5	ug/L	12	Standard
>	**In**	115	686408.4	2.0				ug/L	587441	Standard
	Sn	118	1350.1	6.2	**0.1033**	0.015	14.9	ug/L	497	Standard
	Sb	123	1853.5	25.3	**0.3348**	0.082	24.4	ug/L	78	Standard
	Ba	135	131659.9	0.9	**55.0560**	1.555	2.8	ug/L	24	Standard
	Ce	140	2260.2	0.8				ug/L	25	Standard
>	**Tb**	159	1046624.7	1.4				ug/L	939701	Standard
	Ho	165	75.0	29.1				ug/L	10	Standard
	Tl	203	216.3	9.4	**0.0226**	0.002	8.3	ug/L	4	Standard
	Tl	205	203.3	8.6	**0.0265**	0.002	7.2	ug/L	2	Standard
	Pb	206	535.3	4.4	**0.0217**	0.002	11.2	ug/L	240	Standard
	Pb	207	421.7	4.6	**0.0124**	0.003	25.0	ug/L	198	Standard
	Pb	208	1891.4	2.2	**0.0194**	0.001	4.1	ug/L	883	Standard
	U	238	92.7	15.8	**0.0080**	0.002	21.9	ug/L	3	Standard
>	**Bi**	209	565094.9	1.8				ug/L	523984	Standard

Sample ID: L1606097401 WG573291-01
Report Date/Time: Wednesday, June 29, 2016 20:26:05
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.0814**	27.845	173.2	mg/L	0	Standard
Mg	24	5874.5	4.9	**6.2016**	0.379	6.1	mg/L	37	Standard
K	39	696.7	5.3	**1.1574**	0.052	4.5	mg/L	38	Standard
Ca	43	108.3	19.2	**3.1206**	1.021	32.7	mg/L	10	Standard
Fe	54	752.2	10.4	**0.3885**	0.054	13.9	mg/L	179	Standard
Fe	57	410.0	11.6	**0.0774**	0.139	179.8	mg/L	335	Standard
Sc-1	45	73884.7	1.3				mg/L	61029	Standard
Cl	35	27603.8	2.8				ug/L	20195	Standard
Kr	83	3.0	66.7				ug/L	4	Standard
Br	81	11711.2	8.2				ug/L	1547	Standard
P	31	22032.7	3.4				ug/L	15286	Standard
S	34	2200.2	1.4				ug/L	1572	Standard
Sr	88	218.3	6.6				ug/L	73	Standard
C	12	103.3	43.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	96.2	10.7				mg/L	6	Standard
Ho-1	165	75.0	29.1				mg/L	10	Standard
Er	166	80.0	33.1				mg/L	13	Standard
I	127	173019.2	10.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		109.170	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		110.680	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606097401 WG573291-01
Report Date/Time: Wednesday, June 29, 2016 20:26:05
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.847
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.846
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606097401 WG573291-01
Report Date/Time: Wednesday, June 29, 2016 20:26:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097409S WG573291-05
Sample Date/Time: Wednesday, June 29, 2016 20:27:00
Number of Replicates: 3
Autosampler Position: 413
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	77620.1	3.1				ug/L	74699	Standard
	Be	9	53947.6	2.9	**58.9852**	0.884	1.5	ug/L	8	Standard
	Al	27	10156417.9	2.3	**66.5068**	0.592	0.9	ug/L	1493	Standard
	Sc	45	69663.2	3.3				ug/L	61029	Standard
	Ti	47	217.3	7.1	**0.4124**	0.032	7.7	ug/L	35	Standard
	V	51	458159.6	2.1	**55.0030**	1.779	3.2	ug/L	1435	Standard
	Cr	52	452840.8	0.8	**54.5717**	1.179	2.2	ug/L	7635	Standard
	Cr	53	56488.5	2.6	**55.2901**	2.221	4.0	ug/L	363	Standard
	Mn	55	2082250.4	1.8	**271.9842**	8.205	3.0	ug/L	747	Standard
	Co	59	416796.8	1.5	**57.3937**	1.632	2.8	ug/L	234	Standard
	Ni	60	112326.5	0.8	**54.4621**	1.176	2.2	ug/L	215	Standard
	Cu	65	110417.6	0.6	**53.8701**	0.759	1.4	ug/L	104	Standard
	Zn	66	70959.5	1.2	**58.6468**	1.068	1.8	ug/L	363	Standard
>	Ge	72	544304.5	1.4				ug/L	495142	Standard
	As	75	66583.4	2.1	**54.1492**	1.789	3.3	ug/L	-50	Standard
	Se	82	6605.0	1.3	**54.6200**	1.423	2.6	ug/L	14	Standard
	Se-1	77	4636.4	0.8	**57.1316**	0.988	1.7	ug/L	42	Standard
>	Ga	71	90.0	19.2				mg/L	42	Standard
	Rb	85	30654.3	2.7				ug/L	17	Standard
	Y	89	497057.5	2.9				ug/L	417098	Standard
>	Rh	103	38.3	52.7				ug/L	13	Standard
	Mo	98	163.0	7.4	**0.0487**	0.002	4.2	ug/L	11	Standard
	Ag	107	371099.4	2.0	**50.7567**	1.341	2.6	ug/L	68	Standard
	Cd	111	118377.9	2.8	**50.4636**	1.124	2.2	mg/L	2	Standard
	Cd	114	281427.9	1.4	**48.9324**	1.490	3.0	ug/L	12	Standard
>	In	115	659308.0	4.2				ug/L	587441	Standard
	Sn	118	1090.0	7.1	**0.0721**	0.019	25.7	ug/L	497	Standard
	Sb	123	259620.0	1.8	**49.5761**	1.212	2.4	ug/L	78	Standard
	Ba	135	232712.5	1.7	**101.3728**	3.182	3.1	ug/L	24	Standard
	Ce	140	941.7	8.0				ug/L	25	Standard
>	Tb	159	982549.7	3.5				ug/L	939701	Standard
	Ho	165	31.7	24.1				ug/L	10	Standard
	Tl	203	453507.3	1.5	**53.4353**	0.634	1.2	ug/L	4	Standard
	Tl	205	415041.8	2.5	**56.3402**	0.862	1.5	ug/L	2	Standard
	Pb	206	296774.5	1.1	**53.8916**	0.589	1.1	ug/L	240	Standard
	Pb	207	252251.9	1.7	**51.5506**	0.378	0.7	ug/L	198	Standard
	Pb	208	1071354.8	2.3	**55.4200**	1.065	1.9	ug/L	883	Standard
	U	238	381907.7	2.6	**53.8787**	0.717	1.3	ug/L	3	Standard
>	Bi	209	539522.1	2.0				ug/L	523984	Standard

Sample ID: L1606097409S WG573291-05
Report Date/Time: Wednesday, June 29, 2016 20:29:17
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.4920**	30.288	173.2	mg/L	0	Standard
Mg	24	5339.3	3.4	**5.9734**	0.106	1.8	mg/L	37	Standard
K	39	681.7	10.6	**1.2038**	0.100	8.3	mg/L	38	Standard
Ca	43	128.3	25.3	**4.6217**	2.038	44.1	mg/L	10	Standard
Fe	54	542.4	5.6	**0.2649**	0.028	10.4	mg/L	179	Standard
Fe	57	371.7	15.7	**0.0334**	0.176	526.0	mg/L	335	Standard
> **Sc-1**	45	69663.2	3.3				mg/L	61029	Standard
Cl	35	28661.1	2.0				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	10680.4	5.1				ug/L	1547	Standard
P	31	21703.9	3.0				ug/L	15286	Standard
S	34	2123.5	4.7				ug/L	1572	Standard
Sr	88	153.3	22.2				ug/L	73	Standard
C	12	123.3	28.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	57.6	48.4				mg/L	6	Standard
Ho-1	165	31.7	24.1				mg/L	10	Standard
Er	166	50.0	91.7				mg/L	13	Standard
I	127	154285.8	7.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.911	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.929	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097409S WG573291-05
Report Date/Time: Wednesday, June 29, 2016 20:29:17
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	112.234
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.965
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606097409S WG573291-05
Report Date/Time: Wednesday, June 29, 2016 20:29:17
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097411SD WG573291-06
Sample Date/Time: Wednesday, June 29, 2016 20:30:11
Number of Replicates: 3
Autosampler Position: 414
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77405.6	1.5				ug/L	74699	Standard
	Be	9	55655.4	2.7	**61.0164**	1.576	2.6	ug/L	8	Standard
	Al	27	12086433.4	1.6	**79.3694**	1.985	2.5	ug/L	1493	Standard
	Sc	45	69212.8	2.6				ug/L	61029	Standard
	Ti	47	325.0	3.5	**0.6412**	0.031	4.8	ug/L	35	Standard
	V	51	466207.8	0.7	**53.8111**	1.021	1.9	ug/L	1435	Standard
	Cr	52	460631.7	1.4	**53.3535**	1.026	1.9	ug/L	7635	Standard
	Cr	53	56936.8	2.6	**53.5625**	1.349	2.5	ug/L	363	Standard
	Mn	55	2273187.5	1.8	**285.5286**	8.405	2.9	ug/L	747	Standard
	Co	59	417957.8	1.5	**55.3411**	1.481	2.7	ug/L	234	Standard
	Ni	60	114539.0	0.3	**53.3968**	0.715	1.3	ug/L	215	Standard
	Cu	65	111050.1	0.5	**52.0971**	0.842	1.6	ug/L	104	Standard
	Zn	66	69817.2	1.2	**55.4674**	1.299	2.3	ug/L	363	Standard
>	**Ge**	72	566018.8	1.2				ug/L	495142	Standard
	As	75	67402.1	1.8	**52.7097**	1.486	2.8	ug/L	-50	Standard
	Se	82	6620.3	2.0	**52.6418**	1.659	3.2	ug/L	14	Standard
	Se-1	77	4557.0	3.8	**53.9777**	2.700	5.0	ug/L	42	Standard
>	**Ga**	71	128.3	43.8				mg/L	42	Standard
	Rb	85	32409.6	0.5				ug/L	17	Standard
	Y	89	470382.0	1.0				ug/L	417098	Standard
>	**Rh**	103	61.7	24.8				ug/L	13	Standard
	Mo	98	196.1	5.9	**0.0618**	0.004	5.9	ug/L	11	Standard
	Ag	107	374581.1	0.3	**52.7275**	0.291	0.6	ug/L	68	Standard
	Cd	111	118658.6	1.5	**52.0656**	0.397	0.8	mg/L	2	Standard
	Cd	114	286274.0	2.6	**51.2138**	0.959	1.9	ug/L	12	Standard
>	**In**	115	640194.5	0.7				ug/L	587441	Standard
	Sn	118	1251.7	9.2	**0.1018**	0.017	16.3	ug/L	497	Standard
	Sb	123	263336.7	0.4	**51.7531**	0.327	0.6	ug/L	78	Standard
	Ba	135	250997.3	0.8	**112.5129**	0.934	0.8	ug/L	24	Standard
	Ce	140	2000.1	6.5				ug/L	25	Standard
>	**Tb**	159	984120.1	1.6				ug/L	939701	Standard
	Ho	165	48.3	11.9				ug/L	10	Standard
	Tl	203	459854.1	0.1	**53.6772**	0.823	1.5	ug/L	4	Standard
	Tl	205	414544.5	0.5	**55.7557**	1.089	2.0	ug/L	2	Standard
	Pb	206	298100.5	1.1	**53.6266**	1.109	2.1	ug/L	240	Standard
	Pb	207	253941.0	0.3	**51.4145**	0.875	1.7	ug/L	198	Standard
	Pb	208	1061848.2	0.5	**54.4146**	0.782	1.4	ug/L	883	Standard
	U	238	385197.1	0.6	**53.8449**	1.070	2.0	ug/L	3	Standard
>	**Bi**	209	544631.3	1.4				ug/L	523984	Standard

Sample ID: L1606097411SD WG573291-06
Report Date/Time: Wednesday, June 29, 2016 20:32:28
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	5929.5	2.1	**6.6847**	0.075	1.1	mg/L	37	Standard
K	39	801.7	4.2	**1.4479**	0.027	1.8	mg/L	38	Standard
Ca	43	125.0	31.2	**4.4423**	2.177	49.0	mg/L	10	Standard
Fe	54	841.7	6.4	**0.4916**	0.050	10.2	mg/L	179	Standard
Fe	57	441.7	5.4	**0.2489**	0.040	15.9	mg/L	335	Standard
> **Sc-1**	45	69212.8	2.6				mg/L	61029	Standard
Cl	35	28869.5	0.9				ug/L	20195	Standard
Kr	83	3.7	56.8				ug/L	4	Standard
Br	81	12511.8	0.5				ug/L	1547	Standard
P	31	22511.8	6.4				ug/L	15286	Standard
S	34	2320.2	4.3				ug/L	1572	Standard
Sr	88	170.0	8.8				ug/L	73	Standard
C	12	80.0	25.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	93.8	22.7				mg/L	6	Standard
Ho-1	165	48.3	11.9				mg/L	10	Standard
Er	166	60.0	57.7				mg/L	13	Standard
I	127	186712.7	7.5				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		103.624	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		114.314	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606097411SD WG573291-06
Report Date/Time: Wednesday, June 29, 2016 20:32:28
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	108.980
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	103.940
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606097411SD WG573291-06
Report Date/Time: Wednesday, June 29, 2016 20:32:28
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401PS WG573790-03
Sample Date/Time: Wednesday, June 29, 2016 20:33:23
Number of Replicates: 3
Autosampler Position: 415
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77767.4	1.4				ug/L	74699	Standard
	Be	9	54620.0	3.0	**59.5914**	1.002	1.7	ug/L	8	Standard
	Al	27	11452990.5	1.0	**74.8575**	1.594	2.1	ug/L	1493	Standard
	Sc	45	70666.1	1.7				ug/L	61029	Standard
	Ti	47	337.7	3.5	**0.6996**	0.031	4.4	ug/L	35	Standard
	V	51	448357.3	1.6	**53.6764**	0.056	0.1	ug/L	1435	Standard
	Cr	52	455460.2	1.4	**54.7521**	0.846	1.5	ug/L	7635	Standard
	Cr	53	57375.1	1.7	**56.0052**	0.200	0.4	ug/L	363	Standard
	Mn	55	2154709.7	0.5	**280.7294**	3.122	1.1	ug/L	747	Standard
	Co	59	404569.2	0.4	**55.5661**	0.653	1.2	ug/L	234	Standard
	Ni	60	112633.8	1.0	**54.4757**	0.921	1.7	ug/L	215	Standard
	Cu	65	109355.9	1.1	**53.2239**	1.011	1.9	ug/L	104	Standard
	Zn	66	70498.3	1.6	**58.1166**	0.789	1.4	ug/L	363	Standard
>	**Ge**	72	545627.2	1.6				ug/L	495142	Standard
	As	75	64850.8	0.9	**52.6029**	0.345	0.7	ug/L	-50	Standard
	Se	82	6485.5	1.8	**53.4847**	0.149	0.3	ug/L	14	Standard
	Se-1	77	4411.0	1.9	**54.1848**	0.787	1.5	ug/L	42	Standard
>	**Ga**	71	161.7	1.8				mg/L	42	Standard
	Rb	85	30524.1	1.7				ug/L	17	Standard
	Y	89	474875.4	0.7				ug/L	417098	Standard
>	**Rh**	103	60.0	16.7				ug/L	13	Standard
	Mo	98	163.5	8.5	**0.0491**	0.004	8.3	ug/L	11	Standard
	Ag	107	365805.1	1.5	**50.1795**	1.312	2.6	ug/L	68	Standard
	Cd	111	117534.8	0.9	**50.2525**	0.249	0.5	mg/L	2	Standard
	Cd	114	285376.1	1.5	**49.7475**	0.497	1.0	ug/L	12	Standard
>	**In**	115	657058.7	1.4				ug/L	587441	Standard
	Sn	118	1523.4	8.4	**0.1383**	0.017	12.1	ug/L	497	Standard
	Sb	123	259036.8	1.0	**49.6101**	1.139	2.3	ug/L	78	Standard
	Ba	135	239792.4	1.1	**104.7386**	1.629	1.6	ug/L	24	Standard
	Ce	140	2208.5	2.2				ug/L	25	Standard
>	**Tb**	159	1019758.9	1.4				ug/L	939701	Standard
	Ho	165	71.7	21.3				ug/L	10	Standard
	Tl	203	456925.9	1.0	**51.8248**	0.434	0.8	ug/L	4	Standard
	Tl	205	410087.8	1.3	**53.5969**	1.082	2.0	ug/L	2	Standard
	Pb	206	289464.5	1.3	**50.5998**	1.170	2.3	ug/L	240	Standard
	Pb	207	248071.4	0.9	**48.8073**	1.130	2.3	ug/L	198	Standard
	Pb	208	1050860.8	0.8	**52.3325**	1.144	2.2	ug/L	883	Standard
	U	238	380364.0	0.7	**51.6678**	1.019	2.0	ug/L	3	Standard
>	**Bi**	209	560465.1	1.6				ug/L	523984	Standard

Sample ID: L1606097401PS WG573790-03
Report Date/Time: Wednesday, June 29, 2016 20:35:40
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.2901**	28.207	173.2	mg/L	0	Standard
Mg	24	5812.8	3.2	**6.4153**	0.191	3.0	mg/L	37	Standard
K	39	706.7	5.8	**1.2347**	0.072	5.8	mg/L	38	Standard
Ca	43	108.3	32.4	**3.3848**	1.922	56.8	mg/L	10	Standard
Fe	54	723.0	16.6	**0.3920**	0.096	24.5	mg/L	179	Standard
Fe	57	435.0	5.7	**0.2027**	0.073	35.8	mg/L	335	Standard
> **Sc-1**	45	70666.1	1.7				mg/L	61029	Standard
Cl	35	28625.7	0.6				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	11964.7	4.0				ug/L	1547	Standard
P	31	22167.9	1.3				ug/L	15286	Standard
S	34	2268.5	2.8				ug/L	1572	Standard
Sr	88	201.7	11.5				ug/L	73	Standard
C	12	93.3	34.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	3	Standard
Dy	164	120.1	22.8				mg/L	6	Standard
Ho-1	165	71.7	21.3				mg/L	10	Standard
Er	166	66.7	60.6				mg/L	13	Standard
I	127	179600.9	7.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		104.108	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.196	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097401PS WG573790-03
Report Date/Time: Wednesday, June 29, 2016 20:35:40
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.851
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.962
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606097401PS WG573790-03
Report Date/Time: Wednesday, June 29, 2016 20:35:40
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401SDL WG573790-04
Sample Date/Time: Wednesday, June 29, 2016 20:36:34
Number of Replicates: 3
Autosampler Position: 416
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77070.7	3.6				ug/L	74699	Standard
	Be	9	45.0	33.3	**0.0568**	0.015	26.1	ug/L	8	Standard
	Al	27	2260304.8	1.2	**14.9037**	0.448	3.0	ug/L	1493	Standard
	Sc	45	68278.6	3.1				ug/L	61029	Standard
	Ti	47	83.0	9.9	**0.0892**	0.019	21.8	ug/L	35	Standard
	V	51	2192.5	3.4	**0.0673**	0.008	12.3	ug/L	1435	Standard
	Cr	52	10144.1	1.9	**0.2156**	0.020	9.5	ug/L	7635	Standard
	Cr	53	686.7	4.8	**0.3429**	0.034	10.1	ug/L	363	Standard
	Mn	55	355711.8	1.3	**46.1130**	0.456	1.0	ug/L	747	Standard
	Co	59	2492.2	1.8	**0.3045**	0.007	2.2	ug/L	234	Standard
	Ni	60	828.7	2.9	**0.2769**	0.011	3.8	ug/L	215	Standard
	Cu	65	280.0	0.7	**0.0325**	0.001	2.6	ug/L	104	Standard
	Zn	66	2517.5	2.7	**1.6325**	0.059	3.6	ug/L	363	Standard
>	**Ge**	72	546825.8	0.3				ug/L	495142	Standard
	As	75	65.5	45.7	**0.1126**	0.024	21.6	ug/L	-50	Standard
	Se	82	28.6	15.9	**0.0977**	0.038	38.5	ug/L	14	Standard
	Se-1	77	61.3	12.0	**0.1380**	0.090	65.2	ug/L	42	Standard
>	**Ga**	71	98.3	5.9				mg/L	42	Standard
	Rb	85	6301.3	2.8				ug/L	17	Standard
	Y	89	489576.3	2.2				ug/L	417098	Standard
>	**Rh**	103	21.7	48.0				ug/L	13	Standard
	Mo	98	47.0	21.8	**0.0100**	0.004	35.1	ug/L	11	Standard
	Ag	107	179.0	15.1	**0.0115**	0.004	30.5	ug/L	68	Standard
	Cd	111	35.6	11.4	**0.0072**	0.002	22.2	mg/L	2	Standard
	Cd	114	102.0	28.8	**0.0155**	0.005	31.8	ug/L	12	Standard
>	**In**	115	661161.0	0.8				ug/L	587441	Standard
	Sn	118	795.0	15.7	**0.0261**	0.018	68.8	ug/L	497	Standard
	Sb	123	2322.3	23.7	**0.4382**	0.107	24.4	ug/L	78	Standard
	Ba	135	26230.7	1.3	**11.3705**	0.051	0.4	ug/L	24	Standard
	Ce	140	480.0	13.3				ug/L	25	Standard
>	**Tb**	159	1000122.3	1.4				ug/L	939701	Standard
	Ho	165	26.7	10.8				ug/L	10	Standard
	Tl	203	156.0	22.9	**0.0163**	0.004	24.4	ug/L	4	Standard
	Tl	205	150.0	15.3	**0.0202**	0.003	14.3	ug/L	2	Standard
	Pb	206	403.0	8.6	**0.0007**	0.006	871.4	ug/L	240	Standard
	Pb	207	342.3	5.9	**-0.0013**	0.004	288.2	ug/L	198	Standard
	Pb	208	1497.0	5.1	**0.0020**	0.004	181.5	ug/L	883	Standard
	U	238	114.0	13.2	**0.0113**	0.002	17.4	ug/L	3	Standard
>	**Bi**	209	549941.1	1.0				ug/L	523984	Standard

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:38:51
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	1213.4	3.4	**1.3365**	0.084	6.3	mg/L	37	Standard
K	39	180.0	11.1	**0.2425**	0.050	20.5	mg/L	38	Standard
Ca	43	46.7	22.3	**0.1064**	0.547	514.2	mg/L	10	Standard
Fe	54	286.0	10.0	**0.0782**	0.017	21.6	mg/L	179	Standard
Fe	57	306.7	10.9	**-0.1437**	0.079	55.0	mg/L	335	Standard
> **Sc-1**	45	68278.6	3.1				mg/L	61029	Standard
Cl	35	28506.1	0.3				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	3947.2	6.9				ug/L	1547	Standard
P	31	19479.2	3.3				ug/L	15286	Standard
S	34	2088.5	5.3				ug/L	1572	Standard
Sr	88	101.7	24.8				ug/L	73	Standard
C	12	60.0	60.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	52.5	9.9				mg/L	6	Standard
Ho-1	165	26.7	10.8				mg/L	10	Standard
Er	166	16.7	69.3				mg/L	13	Standard
I	127	39316.0	2.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.175	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.438	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:38:51
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.549
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.954
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401SDL WG573790-04
Sample Date/Time: Wednesday, June 29, 2016 20:39:46
Number of Replicates: 3
Autosampler Position: 417
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76547.9	2.1				ug/L	74699	Standard
	Be	9	28.3	27.0	**0.0389**	0.008	20.5	ug/L	8	Standard
	Al	27	465099.1	0.7	**3.0756**	0.044	1.4	ug/L	1493	Standard
	Sc	45	68169.7	0.9				ug/L	61029	Standard
	Ti	47	36.7	6.9	**-0.0243**	0.006	24.2	ug/L	35	Standard
	V	51	2063.3	3.2	**0.0438**	0.007	16.8	ug/L	1435	Standard
	Cr	52	9662.1	0.8	**0.1188**	0.007	6.3	ug/L	7635	Standard
	Cr	53	523.3	10.2	**0.1664**	0.051	30.7	ug/L	363	Standard
	Mn	55	72938.2	2.1	**9.0325**	0.175	1.9	ug/L	747	Standard
	Co	59	731.7	15.3	**0.0598**	0.015	24.7	ug/L	234	Standard
	Ni	60	309.0	4.2	**0.0207**	0.006	30.5	ug/L	215	Standard
	Cu	65	310.7	4.3	**0.0426**	0.006	14.7	ug/L	104	Standard
	Zn	66	2601.2	2.1	**1.6324**	0.044	2.7	ug/L	363	Standard
>	**Ge**	72	565002.4	0.2				ug/L	495142	Standard
	As	75	-33.8	92.8	**0.0331**	0.025	74.3	ug/L	-50	Standard
	Se	82	20.9	22.0	**0.0285**	0.036	128.0	ug/L	14	Standard
	Se-1	77	57.7	17.5	**0.0697**	0.122	174.5	ug/L	42	Standard
>	**Ga**	71	58.3	9.9				mg/L	42	Standard
	Rb	85	1335.1	7.2				ug/L	17	Standard
	Y	89	490772.8	1.5				ug/L	417098	Standard
>	**Rh**	103	15.0	57.7				ug/L	13	Standard
	Mo	98	24.2	10.4	**0.0025**	0.001	32.4	ug/L	11	Standard
	Ag	107	143.3	39.3	**0.0071**	0.008	112.3	ug/L	68	Standard
	Cd	111	11.6	65.7	**-0.0029**	0.003	116.5	mg/L	2	Standard
	Cd	114	25.4	61.7	**0.0024**	0.003	115.9	ug/L	12	Standard
>	**In**	115	647709.2	0.5				ug/L	587441	Standard
	Sn	118	543.3	5.3	**-0.0103**	0.004	42.8	ug/L	497	Standard
	Sb	123	784.5	24.6	**0.1481**	0.037	24.8	ug/L	78	Standard
	Ba	135	5327.9	0.7	**2.3449**	0.015	0.6	ug/L	24	Standard
	Ce	140	110.0	7.9				ug/L	25	Standard
>	**Tb**	159	986009.2	0.5				ug/L	939701	Standard
	Ho	165	10.0	100.0				ug/L	10	Standard
	Tl	203	50.7	34.9	**0.0041**	0.002	51.6	ug/L	4	Standard
	Tl	205	46.7	71.3	**0.0064**	0.005	70.3	ug/L	2	Standard
	Pb	206	376.3	19.7	**-0.0043**	0.014	323.1	ug/L	240	Standard
	Pb	207	290.7	20.1	**-0.0119**	0.012	103.4	ug/L	198	Standard
	Pb	208	1284.0	11.0	**-0.0091**	0.008	86.4	ug/L	883	Standard
	U	238	44.0	79.9	**0.0017**	0.005	294.8	ug/L	3	Standard
>	**Bi**	209	552760.9	1.2				ug/L	523984	Standard

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:42:03
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	246.7	9.1	**0.2201**	0.028	12.6	mg/L	37	Standard
K	39	50.0	40.0	**-0.0154**	0.039	250.4	mg/L	38	Standard
Ca	43	20.0	43.3	**-1.3952**	0.477	34.2	mg/L	10	Standard
Fe	54	182.5	25.5	**0.0004**	0.036	9653.1	mg/L	179	Standard
Fe	57	256.7	8.8	**-0.2933**	0.068	23.1	mg/L	335	Standard
Sc-1	45	68169.7	0.9				mg/L	61029	Standard
Cl	35	28987.0	1.4				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2360.2	9.1				ug/L	1547	Standard
P	31	19090.4	2.0				ug/L	15286	Standard
S	34	2190.2	0.6				ug/L	1572	Standard
Sr	88	73.3	27.6				ug/L	73	Standard
C	12	46.7	44.6				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	22.7	102.4				mg/L	6	Standard
Ho-1	165	10.0	100.0				mg/L	10	Standard
Er	166	13.3	114.6				mg/L	13	Standard
I	127	10453.6	7.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.475	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.109	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:42:03
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.259
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.492
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 20:42:59
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	73968.6	3.7				ug/L	74699	Standard
	Be	9	51786.8	4.7	**59.3953**	0.618	1.0	ug/L	8	Standard
	Al	27	8138418.2	2.1	**55.9353**	1.016	1.8	ug/L	1493	Standard
	Sc	45	66422.0	2.7				ug/L	61029	Standard
	Ti	47	44367.5	1.4	**105.9251**	2.261	2.1	ug/L	35	Standard
	V	51	448161.7	0.9	**53.5504**	0.563	1.1	ug/L	1435	Standard
	Cr	52	438098.2	0.5	**52.5206**	0.674	1.3	ug/L	7635	Standard
	Cr	53	53648.2	1.8	**52.2574**	1.658	3.2	ug/L	363	Standard
	Mn	55	415062.2	0.8	**53.8560**	1.244	2.3	ug/L	747	Standard
	Co	59	407202.7	0.3	**55.8201**	1.033	1.9	ug/L	234	Standard
	Ni	60	107214.7	0.8	**51.7442**	0.748	1.4	ug/L	215	Standard
	Cu	65	106835.8	1.9	**51.8884**	1.090	2.1	ug/L	104	Standard
	Zn	66	63121.9	0.7	**51.8882**	0.831	1.6	ug/L	363	Standard
>	Ge	72	546729.2	1.7				ug/L	495142	Standard
	As	75	63950.8	1.3	**51.7792**	1.311	2.5	ug/L	-50	Standard
	Se	82	6522.2	2.4	**53.7001**	1.978	3.7	ug/L	14	Standard
	Se-1	77	4462.7	1.0	**54.7224**	0.944	1.7	ug/L	42	Standard
>	Ga	71	48.3	11.9				mg/L	42	Standard
	Rb	85	921.7	10.0				ug/L	17	Standard
	Y	89	469877.0	1.2				ug/L	417098	Standard
>	Rh	103	31.7	45.6				ug/L	13	Standard
	Mo	98	306064.1	0.6	**105.7488**	1.650	1.6	ug/L	11	Standard
	Ag	107	358011.4	0.7	**50.5757**	0.461	0.9	ug/L	68	Standard
	Cd	111	116953.2	1.0	**51.5035**	0.389	0.8	mg/L	2	Standard
	Cd	114	288128.0	0.1	**51.7395**	0.785	1.5	ug/L	12	Standard
>	In	115	637946.0	1.6				ug/L	587441	Standard
	Sn	118	327214.8	2.0	**51.4831**	0.238	0.5	ug/L	497	Standard
	Sb	123	254525.5	0.5	**50.2034**	0.722	1.4	ug/L	78	Standard
	Ba	135	108828.5	1.0	**48.9489**	0.332	0.7	ug/L	24	Standard
	Ce	140	90.0	29.4				ug/L	25	Standard
>	Tb	159	958314.6	1.8				ug/L	939701	Standard
	Ho	165	11.7	49.5				ug/L	10	Standard
	Tl	203	447416.0	0.6	**53.4811**	1.019	1.9	ug/L	4	Standard
	Tl	205	409322.9	1.6	**56.3810**	1.712	3.0	ug/L	2	Standard
	Pb	206	283417.6	0.7	**52.2045**	0.689	1.3	ug/L	240	Standard
	Pb	207	252263.1	0.7	**52.2984**	0.513	1.0	ug/L	198	Standard
	Pb	208	1062273.6	1.1	**55.7538**	1.546	2.8	ug/L	883	Standard
	U	238	385444.7	0.9	**55.1725**	1.082	2.0	ug/L	3	Standard
>	Bi	209	531871.1	1.6				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4377.3	2.2	**5.1285**	0.149	2.9	mg/L	37	Standard
K	39	2180.2	1.8	**4.3148**	0.109	2.5	mg/L	38	Standard
Ca	43	118.3	10.6	**4.3366**	0.588	13.6	mg/L	10	Standard
Fe	54	6990.5	4.3	**5.3088**	0.233	4.4	mg/L	179	Standard
Fe	57	2083.5	2.8	**5.4386**	0.352	6.5	mg/L	335	Standard
> **Sc-1**	45	66422.0	2.7				mg/L	61029	Standard
Cl	35	29309.7	0.2				ug/L	20195	Standard
Kr	83	2.3	24.7				ug/L	4	Standard
Br	81	2156.8	3.9				ug/L	1547	Standard
P	31	21223.2	2.6				ug/L	15286	Standard
S	34	2665.2	2.0				ug/L	1572	Standard
Sr	88	71.7	8.1				ug/L	73	Standard
C	12	56.7	40.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	19.8	1.4				mg/L	6	Standard
Ho-1	165	11.7	49.5				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	2623.6	23.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	118.791		
Al	27	111.871		
Sc	45			
Ti	47	105.925		
V	51	107.101		
Cr	52	105.041		
Cr	53			
Mn	55	107.712		
Co	59	111.640		
Ni	60	103.488		
Cu	65	103.777		
Zn	66	103.776		
> Ge	72		110.419	
As	75	103.558		
Se	82	107.400		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	105.749		
Ag	107	101.151		
Cd	111	103.007		
Cd	114			
In	115		108.597	
Sn	118	102.966		
Sb	123	100.407		
Ba	135	97.898		
Ce	140			
Tb	159			
Ho	165			
Tl	203	106.962		
Tl	205			
Pb	206			
Pb	207			
Pb	208	111.508		
U	238	110.345		
Bi	209		101.505	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 3

Approved: July 04, 2016

QC Std 6	Pb	208
QC Std 6	U	238

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 20:46:11
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	75383.8	1.3				ug/L	74699	Standard
	Be	9	28.3	20.4	**0.0395**	0.006	15.9	ug/L	8	Standard
	Al	27	2638.6	20.0	**0.0044**	0.003	78.6	ug/L	1493	Standard
	Sc	45	68713.9	3.5				ug/L	61029	Standard
	Ti	47	37.7	13.4	**-0.0232**	0.011	49.4	ug/L	35	Standard
	V	51	1533.2	3.6	**-0.0202**	0.008	41.3	ug/L	1435	Standard
	Cr	52	7879.4	2.2	**-0.1054**	0.029	27.9	ug/L	7635	Standard
	Cr	53	448.3	16.3	**0.0885**	0.063	71.1	ug/L	363	Standard
	Mn	55	901.4	5.8	**-0.0363**	0.007	18.9	ug/L	747	Standard
	Co	59	283.3	10.0	**-0.0002**	0.004	1542.3	ug/L	234	Standard
	Ni	60	186.3	2.6	**-0.0379**	0.003	8.3	ug/L	215	Standard
	Cu	65	132.7	3.4	**-0.0421**	0.001	3.4	ug/L	104	Standard
	Zn	66	415.3	5.4	**-0.1263**	0.014	10.9	ug/L	363	Standard
>	Ge	72	573402.3	1.2				ug/L	495142	Standard
	As	75	-34.2	13.3	**0.0332**	0.003	10.5	ug/L	-50	Standard
	Se	82	17.2	23.7	**-0.0025**	0.031	1279.2	ug/L	14	Standard
	Se-1	77	54.3	12.3	**0.0204**	0.082	402.5	ug/L	42	Standard
>	Ga	71	35.0	37.8				mg/L	42	Standard
	Rb	85	30.0	28.9				ug/L	17	Standard
	Y	89	465691.4	1.8				ug/L	417098	Standard
>	Rh	103	6.7	43.3				ug/L	13	Standard
	Mo	98	268.8	25.2	**0.0874**	0.025	29.0	ug/L	11	Standard
	Ag	107	132.7	5.0	**0.0058**	0.001	22.7	ug/L	68	Standard
	Cd	111	15.2	35.4	**-0.0012**	0.003	209.6	mg/L	2	Standard
	Cd	114	58.4	20.0	**0.0084**	0.002	23.7	ug/L	12	Standard
>	In	115	638175.3	2.2				ug/L	587441	Standard
	Sn	118	866.7	15.8	**0.0416**	0.019	45.6	ug/L	497	Standard
	Sb	123	616.5	30.1	**0.1169**	0.034	29.1	ug/L	78	Standard
	Ba	135	50.7	18.3	**0.0068**	0.005	67.0	ug/L	24	Standard
	Ce	140	28.3	27.0				ug/L	25	Standard
>	Tb	159	990005.2	3.0				ug/L	939701	Standard
	Ho	165	6.7	86.6				ug/L	10	Standard
	Tl	203	92.0	24.5	**0.0086**	0.003	32.5	ug/L	4	Standard
	Tl	205	76.7	43.4	**0.0102**	0.005	44.4	ug/L	2	Standard
	Pb	206	305.7	10.6	**-0.0179**	0.007	38.8	ug/L	240	Standard
	Pb	207	274.7	5.9	**-0.0162**	0.004	23.3	ug/L	198	Standard
	Pb	208	1058.7	7.3	**-0.0216**	0.005	23.5	ug/L	883	Standard
	U	238	60.3	25.7	**0.0037**	0.002	61.2	ug/L	3	Standard
>	Bi	209	564302.2	2.4				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:48:28
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	35.0	37.8	**-0.0253**	0.014	54.1	mg/L	37	Standard
K	39	45.0	19.2	**-0.0261**	0.014	54.4	mg/L	38	Standard
Ca	43	26.7	39.0	**-1.0220**	0.608	59.5	mg/L	10	Standard
Fe	54	162.5	17.8	**-0.0163**	0.018	110.8	mg/L	179	Standard
Fe	57	258.3	8.1	**-0.2948**	0.047	15.9	mg/L	335	Standard
Sc-1	45	68713.9	3.5				mg/L	61029	Standard
Cl	35	28935.6	0.2				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	1740.1	12.2				ug/L	1547	Standard
P	31	21446.8	2.8				ug/L	15286	Standard
S	34	2463.5	5.1				ug/L	1572	Standard
Sr	88	71.7	17.6				ug/L	73	Standard
C	12	43.3	13.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	12.9	115.2				mg/L	6	Standard
Ho-1	165	6.7	86.6				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	2023.5	28.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.806	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:48:28
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.636
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.694
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:48:28
Page 3

Approved: July 04, 2016

2.3.3 Metals CVAA Data (Mercury)

2.3.3.1 Summary Data

Microbac

Lab Report #: L16060822
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060822-02	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	18WW03F-061416	**Prep Method:**	7470A	**Prep Date:**	06/16/2016 07:27
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/26/2016 12:23
Workgroup #:	WG573236	**Analyst:**	AMA	**Run Date:**	06/26/2016 13:14
Collect Date:	06/14/2016 07:40	**Dilution:**	1	**File ID:**	M7.062616.131428
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060822
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060822-03	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	18CPTMW04SW-061416	**Prep Method:**	7470A	**Prep Date:**	06/16/2016 07:27
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/26/2016 12:23
Workgroup #:	WG573236	**Analyst:**	AMA	**Run Date:**	06/26/2016 13:16
Collect Date:	06/14/2016 09:00	**Dilution:**	1	**File ID:**	M7.062616.131659
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060822
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.3.3.2 QC Summary

Example Cold Vapor Mercury Calculations
Hydra AA Mercury Analyzer / CETAC M-7600 Quick Trace Mercury Analyzer

1.0 Initial Calibration (ICAL) Parameters
The system performs linear regression from data consisting of a blank and five standards.

2.0 Calculating the concentration (C) of an element in water using data from run log and quantitation report (note:the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (ug/L)	0.1
Vf = Diluted to Volume (mL)	40
Vi = Aliquot Volume (mL)	40
D = Manual dilution factor, if required (10X = 10)	1
Cx = Concentration of element in ppb (ug/L)	0.1

3.0 Calculating the concentration (C) of an element in soil using data from prep log and quantitation report (note: the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Ws} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (ug/L)	0.1
Vf = Diluted to volume (mL)	40
Ws = Aliquot weight (g)	0.6
D = Manual dilution factor	1
Cx = Concentration of element in ug/kg	6.67

4.0 Adjusting the concentration to dry weight:

$$Cdry = \frac{Cx \times 100}{Px}$$

1 Cx = Concentration calculated as received (wet basis)	6.67
Px = Percent solids of sample (%wt)	80
$Cdry$ = Concentration calculated as dry weight (ug/kg)	8.33

8.33 ug/kg = 0.00833 mg/kg

Microbac Laboratories Inc.
Metals Digest Log

Workgroup: WG572915
Analyst: REK
Spike Analyst: REK
Method: 7470A
Run Date: 06/16/2016 07:27
Hotblock Start Temp: 95.2 @ 07:15
Hotblock End Temp: 95 @ 09:15
Instrument: HB7

SOP: ME404 Revison 17
Spike Solution: STD76639
Spike Witness: VC
H2SO4 Lot #: COA18359
40 & 50 ML. DIGESTION TU COA18772
HNO3 Lot #: COA18905
K2S2O8 1:1 Lot #: RGT37046
KMnO4 1:1 Lot #: RGT37082
Mercury Water ICV Lot #: STD76641
HG H2O STDS 10PPM Lot #: STD76647

	SAMPLE #	Type	Matrix	Initial Amount	Final Volume	Spike Amount	Due Date
1	WG572915-02	BLANK	1	40 mL	40 mL		
2	WG572915-03	LCS	1	40 mL	40 mL	4 mL	
3	L16060764-01	SAMP	1	40 mL	40 mL		06/28/16
4	L16060764-02	SAMP	1	40 mL	40 mL		06/28/16
5	L16060764-03	SAMP	1	40 mL	40 mL		06/28/16
6	WG572915-01	REF	1	40 mL	40 mL		
7	L16060764-04	RS01	1	40 mL	40 mL		06/28/16
8	WG572915-04	MS	1	36 mL	40 mL	4 mL	
9	L16060764-05	MS01	1	36 mL	40 mL	4 mL	06/28/16
10	WG572915-05	MSD	1	36 mL	40 mL	4 mL	
11	L16060764-06	SD01	1	36 mL	40 mL	4 mL	06/28/16
12	L16060794-01	SAMP	1	40 mL	40 mL		06/24/16
13	L16060794-02	SAMP	1	40 mL	40 mL		06/24/16
14	L16060794-03	SAMP	1	40 mL	40 mL		06/24/16
15	L16060794-04	SAMP	1	40 mL	40 mL		06/24/16
16	L16060822-02	SAMP	1	40 mL	40 mL		06/24/16
17	L16060822-03	SAMP	1	40 mL	40 mL		06/24/16
18	L16060823-02	SAMP	1	40 mL	40 mL		06/24/16
19	L16060823-03	SAMP	1	40 mL	40 mL		06/24/16
20	L16060823-04	SAMP	1	40 mL	40 mL		06/24/16
21	L16060840-01	SAMP	1	40 mL	40 mL		06/29/16
22	L16060840-02	SAMP	1	40 mL	40 mL		06/29/16
23	L16060840-03	SAMP	1	40 mL	40 mL		06/29/16
24	L16060840-04	SAMP	1	40 mL	40 mL		06/29/16
25	L16060840-05	SAMP	1	40 mL	40 mL		06/29/16
26	L16060840-06	SAMP	1	40 mL	40 mL		06/29/16

Analyst: [signature]

Reviewer: [signature]

* All calibration and check standards are prepared and digested with sample batch following the procedures in section 7.0 of SOP ME404/ME405.

HB_DIG - Modified 07/26/2012
PDF ID: 4809802
Report generated: 06/16/2016 12:16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: CVAA1 Dataset: 062616A.CSV
Analyst1: AMA Analyst2: N/A
Method: 7470A/245.1 SOP: ME404 Rev: 17
Maintenance Log ID:

Calibration Std: STD76647 ICV Std: STD76641 Post Spike: STD76647
ICSA: N/A ICSAB: N/A Int. Std:
CCV: LLCCV: Tuning Sol :
Stannous : RGT36627 Hydroxylamine : RGT36625

Workgroups: 573236

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	M7.062616.121106	WG574322-01	Calibration Point		1		06/26/16 12:11
2	M7.062616.121337	WG574322-02	Calibration Point		1		06/26/16 12:13
3	M7.062616.121609	WG574322-03	Calibration Point		1		06/26/16 12:16
4	M7.062616.121841	WG574322-04	Calibration Point		1		06/26/16 12:18
5	M7.062616.122114	WG574322-05	Calibration Point		1		06/26/16 12:21
6	M7.062616.122348	WG574322-06	Calibration Point		1		06/26/16 12:23
7	M7.062616.122622	WG574322-07	Initial Calibration Verification		1		06/26/16 12:26
8	M7.062616.122853	WG574322-08	Initial Calib Blank		1		06/26/16 12:28
9	M7.062616.123125	WG574322-09	CCV		1		06/26/16 12:31
10	M7.062616.123356	WG574322-10	CCB		1		06/26/16 12:33
11	M7.062616.123627	WG572915-02	Method/Prep Blank	40/40	1		06/26/16 12:36
12	M7.062616.123858	WG572915-03	Laboratory Control S	40/40	1		06/26/16 12:38
13	M7.062616.124129	L16060764-01	MW-23-004	40/40	1		06/26/16 12:41
14	M7.062616.124401	WG573236-01	Post Digestion Spike		1	L16060764-01	06/26/16 12:44
15	M7.062616.124633	L16060764-02	MW-23-002		1		06/26/16 12:46
16	M7.062616.124905	L16060764-03	FD-061316		1		06/26/16 12:49
17	M7.062616.125137	WG572915-05	Matrix Spike Duplica	36/40	1	L16060764-04	06/26/16 12:51
18	M7.062616.125410	L16060764-04	MW-23-005		1	WG572915-01	06/26/16 12:54
19	M7.062616.125643	WG572915-04	Matrix Spike	36/40	1	L16060764-04	06/26/16 12:56
20	M7.062616.125916	L16060794-01	A08-MW06-Y1S2	40/40	1		06/26/16 12:59
21	M7.062616.130149	WG574322-11	CCV		1		06/26/16 13:01
22	M7.062616.130420	WG574322-12	CCB		1		06/26/16 13:04
23	M7.062616.130654	L16060794-02	A08-MW09-Y1S2	40/40	1		06/26/16 13:06
24	M7.062616.130927	L16060794-03	A09-MW02-Y1S2	40/40	1		06/26/16 13:09
25	M7.062616.131158	L16060794-04	A09-MW07-Y1S2	40/40	1		06/26/16 13:11
26	M7.062616.131428	L16060822-02	18WW03F-061416	40/40	1		06/26/16 13:14
27	M7.062616.131659	L16060822-03	18CPTMW04SW-061416	40/40	1		06/26/16 13:16
28	M7.062616.131930	L16060823-02	126F-061416	40/40	1		06/26/16 13:19
29	M7.062616.132202	L16060823-03	C08-061416	40/40	1		06/26/16 13:22
30	M7.062616.132434	L16060823-04	C03-061416	40/40	1		06/26/16 13:24
31	M7.062616.132706	L16060840-01	MW-24-004R	40/40	1		06/26/16 13:27
32	M7.062616.132938	L16060840-02	MW-24-003R	40/40	1		06/26/16 13:29
33	M7.062616.133211	WG574322-13	CCV		1		06/26/16 13:32
34	M7.062616.133441	WG574322-14	CCB		1		06/26/16 13:34

Page: 1 Approved: June 29, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: CVAA1 Dataset: 062616A.CSV

Analyst1: AMA Analyst2: N/A

Method: 7470A/245.1 SOP: ME404 Rev: 17

Maintenance Log ID:

Calibration Std: STD76647 ICV Std: STD76641 Post Spike: STD76647

ICSA: N/A ICSAB: N/A Int. Std:

CCV: LLCCV: Tuning Sol :

Stannous : RGT36627 Hydroxylamine : RGT36625

Workgroups: 573236

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	M7.062616.133714	L16060840-03	FR-061416-24	40/40	1		06/26/16 13:37
36	M7.062616.133947	L16060840-04	MW-24-002R	40/40	1		06/26/16 13:39
37	M7.062616.134220	L16060840-05	MW-24-005	40/40	1		06/26/16 13:42
38	M7.062616.134454	L16060840-06	EB-LOW FLOW TUBING	40/40	1		06/26/16 13:44
39	M7.062616.134726	WG574322-15	CCV		1		06/26/16 13:47
40	M7.062616.134957	WG574322-16	CCB		1		06/26/16 13:49

Page: 2 Approved: June 29, 2016

Microbac Laboratories Inc.

Data Checklist

Date: 26-JUN-2016
Analyst: AMA
Analyst: NA
Method: 7470A/245.1
Instrument: CVAA1
Curve Workgroup: 574322
Runlog ID: 75946
Analytical Workgroups: 573236

STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	
ICB/CCB	X
ICSA/ICSAB	
CRI	
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	0794, 0822, 0823, 0840
Client Forms	X
Level X	
Level 3	0840
Level 4	0794, 0822, 0823
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	AMA
Secondary Reviewer	BKT
Comments	

Primary Reviewer:
28-JUN-2016
Alexandra M. Alfred

Secondary Reviewer:
29-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-29-2016 12:56:19

Microbac Laboratories Inc.

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 7470A

AAB#: WG573236

Login Number: L16060822

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18WW03F-061416	02	06/14/16					06/16/2016	2	28		06/26/16	12.2	28	
18CPTMW04SW-061416	03	06/14/16					06/16/2016	1.9	28		06/26/16	12.2	28	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4828268
Report generated 06/28/2016 10:47

METHOD BLANK SUMMARY

Login Number: L16060822
Blank File ID: M7.062616.123627
Prep Date: 06/16/16 07:27
Analyzed Date: 06/26/16 12:36
Analyst: AMA

Work Group: WG573236
Blank Sample ID: WG572915-02
Instrument ID: CVAA1
Method: 7470A

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG572915-03	M7.062616.123858	06/26/16 12:38	01
18WW03F-061416	L16060822-02	M7.062616.131428	06/26/16 13:14	01
18CPTMW04SW-061416	L16060822-03	M7.062616.131659	06/26/16 13:16	01

Report Name: BLANK_SUMMARY
PDF File ID: 4828269
Report generated 06/28/2016 10:47

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060822 Prep Date: 06/16/16 07:27 Sample ID: WG572915-02
Instrument ID: CVAA1 Run Date: 06/26/16 12:36 Prep Method: 7470A
File ID: M7.062616.123627 Analyst: AMA Method: 7470A
Workgroup (AAB#): WG573236 Matrix: Water Units: mg/L
Contract #: Cal ID: CVAA1-26-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Mercury	0.000100	0.000400	0.000100	1	U

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name:BLANK
PDF ID: 4828270
28-JUN-2016 10:48

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060822 Run Date: 06/26/2016 Sample ID: WG572915-03
Instrument ID: CVAA1 Run Time: 12:38 Prep Method: 7470A
File ID: M7.062616.123858 Analyst: AMA Method: 7470A
Workgroup (AAB#): WG573236 Matrix: Water Units: mg/L
QC Key: DOD4 Lot#: STD76639 Cal ID: CVAA1-26-JUN-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Mercury	0.00400	0.00421	105	80 - 120	

LCS - Modified 03/06/2008
PDF File ID: 4828271
Report generated: 06/28/2016 10:48

Microbac Laboratories Inc.
MATRIX SPIKE AND MATRIX SPIKE DUP (MS/MSD)

Loginnum: L16060822 Cal ID: CVAA1 - Worknum: WG573236
Instrument ID: CVAA1 Contract #: Method: 7470A
Parent ID: WG572915-01 File ID: M7.062616.125410 Dil: 1 Matrix: WATER
Sample ID: WG572915-04 MS File ID: M7.062616.125643 Dil: 1 Units: mg/L
Sample ID: WG572915-05 MSD File ID: M7.062616.125137 Dil: 1

Analyte	Parent	MS Spiked	MS Found	MS %Rec	MSD Spiked	MSD Found	MSD %Rec	%RPD	%Rec Limits	RPD Limit	Q
Mercury	ND	0.00444	0.00479	108	0.00444	0.00480	108	0.0463	80 - 120	20	

* FAILS %REC LIMIT

FAILS RPD LIMIT

NOTE: This is an internal quality control sample.

WG_MS_MSD_DRYWT - Modified 05/26/2011
PDF File ID: 4828272
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
POST SPIKE REPORT

Sample Login ID: L16060822 **Worknum:** WG573236
Instrument ID: CVAA1 **Method:** 7470A
Post Spike ID: WG573236-01 **File ID:** M7.062616.124401 **Dil:** 1 **Units:** ug/L
Sample ID: L16060764-01 **File ID:** M7.062616.124129 **Dil:** 1 **Matrix:** Water

Analyte	Post Spike Result	C	Sample Result	C	Spike Added(SA)	% R	Control Limit %R	Q
MERCURY	1.07		0	U	1	107.1	85 - 115	

N = % Recovery exceeds control limits
F = Result is between MDL and RL
U = Sample result is below MDL. A value of zero is used in the calculation

POST_SPIKE - Modified 03/06/2008
PDF File ID: 4828267
Report generated: 06/28/2016 10:47

Microbac Laboratories Inc.

INITIAL CALIBRATION SUMMARY

Login Number: L16060822
Analytical Method: 7470A
ICAL Worknum: WG574322

Workgroup (AAB#): WG573236
Instrument ID: CVAA1
Initial Calibration Date: 06/26/2016 12:23

Analyte	WG574322-01		WG574322-02		WG574322-03		WG574322-04		WG574322-05		WG574322-06	
	STD	INT	STD	INT	STD	INT	STD	INT	STD	INT	STD	INT
Mercury	0	57.64	0.200	2673	1.00	13280	2.00	26640	5.00	66160	10.0	128600

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
* = Out of Compliance; R < 0.995

INT_CAL_HG_FU - Modified 03/06/2008
PDF File ID: 4828273
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.

INITIAL CALIBRATION SUMMARY

Login Number: L16060822
Analytical Method: 7470A
ICAL Worknum: WG574322

Workgroup (AAB#): WG573236
Instrument ID: CVAA1
Initial Calibration Date: 06/26/2016 12:23

Analyte	R	Q
Mercury	1.000	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
* = Out of Compliance; R < 0.995

INT_CAL_HG_FU - Modified 03/06/2008
PDF File ID: 4828273
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16060822
Run Date: 06/26/2016
Sample ID: WG574322-08
Instrument ID: CVAA1
Run Time: 12:28
Method: 7470A
File ID: M7.062616.122853
Analyst: AMA
Units: ug/L
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER

Analytes	MDL	RDL	Concentration	Qualifier
MERCURY	.1	.4	.1	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4828275
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/26/2016
Sample ID: WG574322-10
Instrument ID: CVAA1
Run Time: 12:33
Method: 7470A
File ID: M7.062616.123356
Analyst: AMA
Units: ug/L
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Mercury	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828277
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/26/2016
Sample ID: WG574322-12
Instrument ID: CVAA1
Run Time: 13:04
Method: 7470A
File ID: M7.062616.130420
Analyst: AMA
Units: ug/L
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Mercury	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828277
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060822
Run Date: 06/26/2016
Sample ID: WG574322-14
Instrument ID: CVAA1
Run Time: 13:34
Method: 7470A
File ID: M7.062616.133441
Analyst: AMA
Units: ug/L
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Mercury	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828277
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.

INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16060822
Run Date: 06/26/2016
Sample ID: WG574322-07
Instrument ID: CVAA1
Run Time: 12:26
Method: 7470A
File ID: M7.062616.122622
Analyst: AMA
Units: ug/L
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Mercury		2	2.12	106	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4828274
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822
Run Date: 06/26/2016
Sample ID: WG574322-09
Instrument ID: CVAA1
Run Time: 12:31
Method: 7470A
File ID: M7.062616.123125
Analyst: AMA
QC Key: DOD4
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Mercury, Total		0.00200	0.00210	mg/L	105	80 - 120		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828276
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822
Run Date: 06/26/2016
Sample ID: WG574322-11
Instrument ID: CVAA1
Run Time: 13:01
Method: 7470A
File ID: M7.062616.130149
Analyst: AMA
QC Key: DOD4
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Mercury, Total		0.00200	0.00219	mg/L	109	80 - 120		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828276
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060822
Run Date: 06/26/2016
Sample ID: WG574322-13
Instrument ID: CVAA1
Run Time: 13:32
Method: 7470A
File ID: M7.062616.133211
Analyst: AMA
QC Key: DOD4
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Mercury, Total		0.00200	0.00223	mg/L	112	80 - 120		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828276
Report generated 06/28/2016 10:48

2.3.3.3 Raw Data

PDM CVAA1 245.1/7470/7471

Report Generated By CETAC QuickTrace
Analyst: VOA
Worksheet file: C:\Program Files (x86)\QuickTrace\Worksheets\062616A.wsz
Date Started: 6/26/2016 11:26:47 AM
Comment:

Results

Sample Name					Type	Date/Time	Conc (ug/L)	µAbs	%RSD	Flags	DF
Standard #0					STD	06/26/16 12:11:06 pm	0.0000	58	36.36		1.00
Replicates	72.2	26.9	61.9	69.6							
Standard #1 (0.2 ug/L)					STD	06/26/16 12:13:37 pm	0.2000	2673	0.35		1.00
Replicates	2686.9	2668.0	2667.5	2669.0							
Standard #2 (1.0 ug/L)					STD	06/26/16 12:16:09 pm	1.0000	13283	0.12		1.00
Replicates	13271.5	13294.2	13266.7	13300.2							
Standard #3 (2.0 ug/L)					STD	06/26/16 12:18:41 pm	2.0000	26641	0.01		1.00
Replicates	26642.8	26641.1	26641.1	26638.4							
Standard #4 (5.0 ug/L)					STD	06/26/16 12:21:14 pm	5.0000	66159	0.11		1.00
Replicates	66221.6	66218.8	66109.5	66085.3							
Standard #5 (10.0 ug/L)					STD	06/26/16 12:23:48 pm	10.0000	128636	0.07		1.00
Replicates	128716.2	128712.3	128583.1	128533.8							

Calibration

Equation: A = 499.952 + 12881.860C
R2: 0.99980
SEE: 796.4388
Flags:



ICV					ICV	06/26/16 12:26:22 pm	2.1170	27772	0.16		1.00
Replicates	27815.3	27806.0	27744.4	27723.2							
% Recovery	105.86										

Approved: June 28, 2016
Alexandra M. Alfred

Sample Name					Type	Date/Time	Conc (ug/L)	µAbs	%RSD	Flags	DF
ICB					ICB	06/26/16 12:28:53 pm	-0.0379	11	291.32		1.00
Replicates	4.1	-22.5	57.6	6.8							
CCV					CCV	06/26/16 12:31:25 pm	2.0950	27485	0.16		1.00
Replicates	27434.8	27461.4	27527.1	27517.9							
% Recovery	104.74										
CCB					CCB	06/26/16 12:33:56 pm	-0.0387	1	964.00		1.00
Replicates	9.0	5.0	-0.8	-9.8							
WG572915-02					UNK	06/26/16 12:36:27 pm	-0.0350	50	39.52		1.00
Replicates	72.8	57.2	40.4	27.9							
WG572915-03					UNK	06/26/16 12:38:58 pm	4.2070	54690	0.15		1.00
Replicates	54575.2	54727.8	54702.1	54755.0							
L1606076401					UNK	06/26/16 12:41:29 pm	-0.0379	11	72.00		1.00
Replicates	8.4	14.3	21.3	1.9							
WG573236-01					SPK	06/26/16 12:44:01 pm	1.0710	14290	0.11		1.00
Replicates	14266.8	14293.4	14303.5	14296.9							
% Recovery	110.84										
L1606076402					UNK	06/26/16 12:46:33 pm	-0.0345	56	29.18		1.00
Replicates	46.7	78.1	41.1	58.6							
L1606076403					UNK	06/26/16 12:49:05 pm	-0.0307	104	20.90		1.00
Replicates	76.8	97.6	116.7	125.9							
WG572915-01					UNK	06/26/16 12:51:37 pm	4.3160	56092	0.13		1.00
Replicates	56018.1	56175.3	56135.9	56040.3							
L1606076405					MSK	06/26/16 12:54:10 pm	-0.0402	-17	79.85	N	1.00
Replicates	-11.0	-38.2	-11.2	-9.2							
% Recovery	-108.89										
L1606076406					MSDUP	06/26/16 12:56:43 pm	4.3140	56069	0.23	D	1.00
Replicates	55903.2	56042.6	56133.8	56194.6							
% Recovery	-0.05	RPD 203.76									

Approved: June 28, 2016
Alexandra M. Alfred

Sample Name					Type	Date/Time	Conc (ug/L)	µAbs	%RSD	Flags	DF
L1606079401					UNK	06/26/16 12:59:16 pm	-0.0393	-7	153.33		1.00
Replicates	-17.5	2.6	1.3	-12.4							
CCV					CCV	06/26/16 01:01:49 pm	2.1870	28674	0.28		1.00
Replicates	28565.4	28662.0	28726.3	28743.6							
% Recovery	109.36										
CCB					CCB	06/26/16 01:04:20 pm	-0.0400	-15	71.00		1.00
Replicates	-20.6	-24.4	-16.6	0.3							
L1606079402					UNK	06/26/16 01:06:54 pm	-0.0360	36	40.24		1.00
Replicates	40.1	23.8	25.1	54.5							
L1606079403					UNK	06/26/16 01:09:27 pm	0.0214	775	1.64		1.00
Replicates	766.0	779.2	763.7	791.0							
L1606079404					UNK	06/26/16 01:11:58 pm	-0.0321	87	18.96		1.00
Replicates	68.6	84.0	86.7	108.7							
L1606082202					UNK	06/26/16 01:14:28 pm	-0.0340	62	33.50		1.00
Replicates	78.3	50.1	79.4	38.2							
L1606082203					UNK	06/26/16 01:16:59 pm	-0.0350	49	26.85		1.00
Replicates	46.5	64.4	52.2	32.7							
L1606082302					UNK	06/26/16 01:19:30 pm	-0.0247	181	8.96		1.00
Replicates	178.7	199.1	187.1	160.4							
L1606082303					UNK	06/26/16 01:22:02 pm	-0.0283	135	14.85		1.00
Replicates	148.4	150.3	135.4	106.8							
L1606082304					UNK	06/26/16 01:24:34 pm	-0.0349	51	26.20		1.00
Replicates	38.5	67.5	41.8	55.0							
L1606084001					UNK	06/26/16 01:27:06 pm	-0.0336	67	31.06		1.00
Replicates	78.3	84.6	37.9	65.6							

Approved: June 28, 2016
Alexandra M. Alfred

Sample Name					Type	Date/Time	Conc (ug/L)	µAbs	%RSD	Flags	DF
L1606084002					UNK	06/26/16 01:29:38 pm	-0.0355	42	37.05		1.00
Replicates	37.5	51.9	22.2	56.9							
CCV					CCV	06/26/16 01:32:11 pm	2.2310	29245	0.47		1.00
Replicates	29065.4	29210.7	29333.0	29372.9							
% Recovery	111.57										
CCB					CCB	06/26/16 01:34:41 pm	-0.0387	1	222.94		1.00
Replicates	7.9	10.6	18.4	-34.1							
L1606084003					UNK	06/26/16 01:37:14 pm	-0.0343	58	10.23		1.00
Replicates	58.3	56.5	65.5	51.2							
L1606084004					UNK	06/26/16 01:39:47 pm	-0.0311	99	14.17		1.00
Replicates	85.9	113.0	88.0	109.0							
L1606084005					UNK	06/26/16 01:42:20 pm	-0.0347	53	38.40		1.00
Replicates	51.7	58.0	26.5	75.7							
L1606084006					UNK	06/26/16 01:44:54 pm	-0.0362	34	52.68		1.00
Replicates	39.2	7.6	40.3	47.5							
CCV					CCV	06/26/16 01:47:26 pm	2.2480	29454	0.57		1.00
Replicates	29236.8	29421.4	29529.4	29626.5							
% Recovery	112.38										
CCB					CCB	06/26/16 01:49:57 pm	-0.0385	4	182.59		1.00
Replicates	14.9	-3.5	0.5	5.6							

Approved: June 28, 2016

Alexandra M. Alfred

3.0 Attachments

Microbac Laboratories Inc.
Ohio Valley Division Analyst List
July 7, 2016

001 - BIO-CHEM TESTING WVDEP 220
002 - REIC Consultants, Inc. WVDEP 060
003 - Sturm Environmental
004 - MICROBAC PITTSBURGH
005 - ES LABORATORIES
006 - ALCOSAN LABORATORIES
007 - ALS LABORATORIES
008 - BENCHMARK LABORATORIES
010 - MICROBAC CHICAGOLAND
AC - AMBER R. CARMICHAEL
ADC - ANTHONY D. CANTER
ADG - APRIL D. GREENE
AED - ALLEN E. DAVIS
ALS - ADRIANE L. STEED
AMA - ALEXANDRA M. ALFRED
AWE - ANDREW W. ESSIG
AZH - AFTER HOURS
BJO - BRIAN J. OGDEN
BKT - BRENDAN TORRENCE
BLG - BRENDA L. GREENWALT
BNB - Brandi N. Bentley
BRG - BRENDA R. GREGORY
CAA - CASSIE A. AUGENSTEIN
CAF - CHERYL A. FLOWERS
CEB - CHAD E. BARNES
CJQ - Cameron J. Quick
CJR - COURTNEY J. REXROAD
CLC - CHRYS L. CRAWFORD
CLS - CARA L. STRICKLER
CLW - CHARISSA L. WINTERS
CPD - CHAD P. DAVIS
CSH - CHRIS S. HILL
CV - Carl Volkman
DAK - DEAN A. KETELSEN
DCM - DAVID C. MERCKLE
DEV - DAVID E. VANDENBERG
DIH - DEANNA I. HESSON
DLB - DAVID L. BUMGARNER
DLP - DOROTHY L. PAYNE
DLW - DIANA L. WRIGHT
DSM - DAVID S. MOSSOR
ECL - ERIC C. LAWSON
EMW - ERIC M. WILKEN
ENY - EMILY N. YOAK
ERP - ERIN R. PORTER
FJB - FRANCES J. BOLDEN
JBK - JEREMY B. KINNEY
JDH - JUSTIN D. HESSON
JDS - JARED D. SMITH
JJS - JOHN J. STE MARIE
JKP - JACQUELINE K. PARSONS
JLD - JESSICA L. DELONG
JLL - JOHN L. LENT
JMW - JEANA M. WHITE
JTP - JOSHUA T. PEMBERTON
JWR - JOHN W. RICHARDS
JWS - JACK W. SHEAVES
JYH - JI Y. HU
KAJ - KELLIE A. JOHNSON
KAT - KATHY A. TUCKER
KDW - KATHRYN D. WELCH
KEB - KATIE E. BARNES
KHR - KIM H. RHODES
KKB - KERRI K. BUCK
KRA - KATHY R. ALBERTSON
KRB - KAELY R. BECKER
KRP - KATHY R. PARSONS
LEC - LAURA E. CARPENTER
LJH - Lacey J. Hendershot
LKN - LINDA K. NEDEFF
LLS - LARRY L. STEPHENS
LSB - LESLIE S. BUCINA
MAP - MARLA A. PORTER
MBK - MORGAN B. KNOWLTON
MDA - MIKE D. ALBERTSON
MDC - MIKE D. COCHRAN
MES - MARY E. SCHILLING
MLB - MEGAN L. BACHE
MMB - MAREN M. BEERY
MRT - MICHELLE R. TAYLOR
MSW - MATT S. WILSON
NPH - Natalie P. Hart
PDM - PIERCE D. MORRIS
PIT - MICROBAC WARRENDALE
PRL - PAIGE R. LAMB
PSW - PEGGY S. WEBB
QX - QIN XU
RAH - ROY A. HALSTEAD
REK - BOB E. KYER
RLB - BOB BUCHANAN
RM - RAYMOND MALEKE
RNP - RICK N. PETTY
RST - ROBIN S. TURNER
SAV - SARAH A. VANDENBERG
SCB - SARAH C. BOGOLIN
SDC - SHALYN D. CONLEY
SLM - STEPHANIE L. MOSSBURG
SLP - SHERI L. PFALZGRAF
TB - TODD BOYLE
TGF - TIM G. FELTON
TMB - TIFFANY M. BAILEY
TMM - TAMMY M. MORRIS
VC - VICKI COLLIER
WJB - WILL J. BEASLEY
WRR - WESLEY R. RICHARDS
WTD - WADE T. DELONG
XXX - UNAVAILABLE OR SUBCONTRACT

Microbac Laboratories Inc.
List of Valid Qualifiers
July 07, 2016

Qualkey: DOD

Qualifier	Description
*	Surrogate or spike compound out of range
+	Correlation coefficient for the MSA is less than 0.995
<	Result is less than the associated numerical value.
>	Greater than
A	See the report narrative
B	The reported result is associated with a contaminated method blank.
B,H1	Analyte present in method blank. Sample analysis performed past holding time.
B1	Target analyte detected in method blank at or above the method reporting limit
B3	Target analyte detected in calibration blank at or above the method reporting limit
B4	The BOD unseeded dilution water blank exceeded 0.2 mg/L
C	Confirmed by GC/MS
CG	Confluent growth
CT1	Cooler temperature at sample reciept exceeded regulatory limit.
DL	Surrogate or spike compound was diluted out
E	Estimated concentration due to sample matrix interference
E,CT1	Estimated results. The cooler temperature at receipt exceeded regulatory guidelines for requested testing.
EDL	Elevated sample reporting limits, presence of non-target analytes
EMPC	Estimated Maximum Possible Concentration
F, S	Estimated result below quantitation limit; method of standard additions(MSA)
F,CT1	Estimated value; the analyte concentration was less than the RL/LOQ. The cooler temperature at receipt exceeded regula
FL	Free Liquid
FP1	Did not ignite.
H1	Sample analysis performed past holding time.
H1,CT1	Sample analysis performed past holding time. The cooler temperature at receipt exceeded regulatory guidelines for reque
I	Semiquantitative result (out of instrument calibration range)
J	Estimated concentration; sample matrix interference.
J	Estimated value ; the analyte concentration was greater than the highest standard
J	Estimated value ; the analyte concentration was less than the LOQ.
J	The reported result is an estimated value.
J,B	Analyte detected in both the method blank and sample above the MDL.
J,CT1	Estimated value ; the analyte concentration was less than the LOQ. Cooler temperature at sample reciept exceeded regu
J,H1	Estimated value ; the analyte concentration was less than the LOQ. Sample analysis performed past holding time.
J,H1	The reported result is an estimated value. Sample was analyzed past holding time.
J,P	Estimate; columns don't agree to within 40%
J,S	Estimated concentration; analyzed by method of standard addition (MSA)
JB	The reported result is an estimated value. The reported result is also associated with a contaminated method blank.
JQ	The reported result is an estimated value and one or more quality control criteria failed. See narrative.
L	Sample reporting limits elevated due to matrix interference
L1	The associated blank spike (LCS) recovery was above the laboratory acceptance limits.
L2	The associated blank spike (LCS) recovery was below the laboratory acceptance limits.
M	Matrix effect; the concentration is an estimate due to matrix effect.
N	Nontarget analyte; the analyte is a tentativlely identified compound (TIC) by GC/MS
NA	Not applicable
ND	Not detected at or above the reporting limit (RL)
ND, B	Not detected at or above the reporting limit (RL). Analyte present in method blank.
ND, CT1	Analyte was not detected. The concentration is below the reported LOD. The cooler temperature at receipt exceeded reg
ND, L	Not detected; sample reporting limit (RL) elevated due to interference
ND, S	Not detected; analyzed by method of standard addition (MSA)
ND,H1	Not detected; Sample analysis performed past holding time.
ND,H1,CT1	Not detected; Sample analysis performed past holding time. The cooler temperature at receipt exceeded regulatory guide
NF	Not found by library search
NFL	No free liquid
NI	Non-ignitable
NR	Analyte is not required to be analyzed
NS	Not spiked
P	Concentrations >40% difference between the two GC columns
Q	One or more quality control criteria failed. See narrative.
Q,H1	One or more quality control criteria failed. Sample analyzed past holding time. See narrative.
QNS	Quantity of sample not sufficient to perform analysis
RA	Reanalysis confirms reported results
RE	Reanalysis confirms sample matrix interference
S	Analyzed by method of standard addition (MSA)
SMI	Sample matrix interference on surrogate
SP	Reported results are for spike compounds only
T5	Laboratory not licensed for this parameter
TIC	Library Search Compound
TNTC	Too numerous to count

Microbac Laboratories Inc.
List of Valid Qualifiers
July 07, 2016

Qualkey: DOD

TNTC, B	Too numerous to count. Analyte present in method blank.
TNTC,CT1	Too numerous to count. The cooler temperature at receipt exceeded regulatory guidelines for requested testing.
TNTC,H1	Too numerous to count. Sample analysis performed past holding time.
U	Analyte was not detected. The concentration is below the reported LOD.
U,CT1	Analyte was not detected. The concentration is below the reported LOD. Cooler temperature at sample reciept exceeded
U,H1	Not detected; Sample analysis performed past holding time.
UJ	Undetected; the MDL and RL are estimated due to quality control discrepancies.
UQ	Undetected; the analyte was analyzed for, but not detected.
W	Post-digestion spike for furnace AA out of control limits
X	Exceeds regulatory limit
X, S	Exceeds regulatory limit; method of standard additions (MSA)
Z	Cannot be resolved from isomer - see below

AECOM

Chain of Custody Record

COC Number:

Laboratory: Microbac POC: Stephanie Mossburg	Project Manager: Debra Richmann	Mail to: Linda Raabe	
Address: 158 Starlite Drive	Phone/Fax Number: 210-296-2000	112 East Pecan STE. 400	
Marietta, OH 45750	Sampler (print): Scott Beesinger	San Antonio, TX 78205	
Phone: 1-800-373-4071		210-296-2000	
Client: AECOM	Signature: Scott Beesinger	Fed Ex Airbill No:	
Address: 112 East Pecan Ste. 400			
San Antonio, TX 78205			
Turn Around Time: STANDARD	pH:	Program:	
Project Name/Location: Longhorn			
Project Number: 60274185.0013BH 1824 MR10		ERPIMS REQUIRED FIELDS	

Site Name	Sample ID/Location ID	SBD	SED	Date	Time	Comp*	Grab	Matrix	Number of Containers	VOC	Perchlorate	Total Metals	1,4 - DIOXANE					SA CODE	Cooler ID	LOT CONTROL NUMBERS ABLOT	EBLOT	TBLOT
SITE 18/24	18WW03-061416			6/14/16	0740		✓	W	6	✓	✓		✓									
	18WW03F-061416			6/14/16	0740		✓	W	1			✓										
	18CPTMW04SW-061416			6/14/16	0900		✓	W	7	✓	✓	✓	✓									
	18CPTMW04-061416			6/14/16	1000		✓	W	6	✓	✓		✓									
	18CPTMW04FD-061416			6/14/16	1000		✓	W	6	✓	✓		✓									
	TRIP BLANK			6/14/16			✓	W	2	✓												

Comments: **STANDARD TAT**

	Date	Time	Received by: (Signature)	Received by: (Signature)	Date	Time	Relinquished by: (Signature)
Relinquished by: (Signature) Scott Beesinger	6/14/16	1530					
Relinquished by: (Signature)	Date	Time	Rec (Sig			Time	Remarks:

Microbac OVD
Received: 06/15/2016 09:54
By: BRENDA GREGORY 221000087198

Brenda Gregory

*Homogenize all composite samples prior to analysis

...nager, Pink QA/QC Manager

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060822
Account: 2551
Project: 2551.096
Samples: 6
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060822-01	760842	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	ORG4	15-JUN-2016 14:47	BRG		
2	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Samplenum	Container ID	Products
L16060822-01	760843	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	PREP	W1	EXT	15-JUN-2016 17:32	CPD	AZH	
3	DISP	EXT	DISP	16-JUN-2016 15:22	BJO	BJO	
4	ANALYZ*	EXT	SEMI	20-JUN-2016 09:18	NPH	CPD	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	STORE	W1	A1	24-JUN-2016 10:00	BRG	BRG	

Sample extract/digestate/leachate

Samplenum	Container ID	Products
L16060822-01	760844	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	ANALYZ	W1	SEM	22-JUN-2016 14:47	JWR	CLS	
3	STORE	SEM	A1	24-JUN-2016 12:04	BRG	JWR	

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060822
Account: 2551
Project: 2551.096
Samples: 6
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060822-02	760845	AG-MS AL AS-MS BA-MS BE-AX CA CD-MS CO-MS CR-M

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	PREP	W1	DIG	15-JUN-2016 14:59	AC	CLS	
3	ANALYZ*	DIG	METALS	17-JUN-2016 14:30	CJQ	AC	
4	STORE	DIG	A1	17-JUN-2016 14:50	CLS	ERP	

Sample extract/digestate/leachate

Samplenum	Container ID	Products
L16060822-03	760846	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:14	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:14	CLS	JST	

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060822
Account: 2551
Project: 2551.096
Samples: 6
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060822-03	760847	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	PREP	W1	EXT	15-JUN-2016 17:32	CPD	AZH	
3	DISP	EXT	DISP	16-JUN-2016 15:23	BJO	BJO	
4	ANALYZ*	EXT	SEMI	20-JUN-2016 09:18	NPH	CPD	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	STORE	W1	A1	24-JUN-2016 08:22	BRG	BRG	

Sample extract/digestate/leachate

Samplenum	Container ID	Products
L16060822-03	760848	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	ANALYZ	W1	SEM	22-JUN-2016 14:47	JWR	CLS	
3	STORE	SEM	A1	24-JUN-2016 12:04	BRG	JWR	

Samplenum	Container ID	Products
L16060822-03	760849	AG-MS AL AS-MS BA-MS BE-AX CA CD-MS CO-MS CR-M

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	PREP	W1	DIG	15-JUN-2016 14:59	AC	CLS	
3	ANALYZ*	DIG	METALS	17-JUN-2016 14:30	CJQ	AC	
4	STORE	DIG	A1	17-JUN-2016 14:50	CLS	ERP	

Sample extract/digestate/leachate

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060822
Account: 2551
Project: 2551.096
Samples: 6
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060822-04	760850	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Samplenum	Container ID	Products
L16060822-04	760851	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	PREP	W1	EXT	15-JUN-2016 17:32	CPD	AZH	
3	DISP	EXT	DISP	16-JUN-2016 15:22	BJO	BJO	
4	ANALYZ*	EXT	SEMI	20-JUN-2016 09:18	NPH	CPD	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	STORE	W1	A1	24-JUN-2016 08:23	BRG	BRG	

Sample extract/digestate/leachate

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060822
Account: 2551
Project: 2551.096
Samples: 6
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060822-04	760852	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	ANALYZ	W1	SEM	22-JUN-2016 14:47	JWR	CLS	
3	STORE	SEM	A1	24-JUN-2016 12:04	BRG	JWR	

Samplenum	Container ID	Products
L16060822-05	760853	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060822
Account: 2551
Project: 2551.096
Samples: 6
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060822-05	760854	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	PREP	W1	EXT	15-JUN-2016 17:32	CPD	AZH	
3	DISP	EXT	DISP	16-JUN-2016 15:22	BJO	BJO	
4	ANALYZ*	EXT	SEMI	20-JUN-2016 09:18	NPH	CPD	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	STORE	W1	A1	24-JUN-2016 08:23	BRG	BRG	

Sample extract/digestate/leachate

Samplenum	Container ID	Products
L16060822-05	760855	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:47	BRG		
2	ANALYZ	W1	SEM	22-JUN-2016 14:47	JWR	CLS	
3	STORE	SEM	A1	24-JUN-2016 12:04	BRG	JWR	

Samplenum	Container ID	Products
L16060822-06	760856	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:47	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

NELAP Addendum - January 4, 2016

Non-NELAP LIMS Product and Description

The following is a list of those tests that are not included in the Microbac – OVD NELAP Scope of Accreditation:

Heat of Combustion (BTU)
Total Halide by Bomb Combustion (TX)
Particle Sizing - 200 Mesh (PS200)
Specific Gravity/Density (SPGRAV)
Total Residual Chlorine (CL-TRL)
Total Volatile Solids (all forms) (TVS)
Total Coliform Bacteria (all methods)
Fecal Coliform Bacteria (all methods)
Sulfite (SO3)
Propionaldehyde (HPLC-UV)

SOLID AND HAZARDOUS CHEMICALS
Nitrogen, Ammonia by Method 350.1
Chromium, Hexavalent, Leachable by SM3500 Cr-B 2009
Phenolics, Total by Method 420.1
ASTM D3987-06

NELAP Accreditation by Laboratory SOP

NONPOTABLE WATER

OVD HPLC02/HPLC-UV

Nitroglycerin
Acetic acid
Butyric acid
Lactic acid
Propionic acid
Pyruvic acid

OVD MSS01/GC-MS

1,4-Phenylenediamine
1-Methylnaphthalene
1,4-Dioxane
Atrazine
Benzaldehyde
Biphenyl
Caprolactam
Hexamethylphosphoramide (HMPA)
Pentachlorobenzene
Pentachloroethane

NELAP Accreditation by Laboratory SOP

NONPOTABLE WATER

OVD MSV01/GC-MS

1, 1, 2-Trichloro-1,2,2-trifluoroethane
1,3-Butadiene
Cyclohexane
Cyclohexanone
Dimethyl disulfide
Dimethylsulfide
Ethyl-t-butylether (ETBE)
Isoprene
Methylacetate
Methylcyclohexane
T-amylmethylether (TAME)
Tetrahydrofuran (THF)

OVD HPLC07/HPLC-MS-MS

Hexamethylphosphoramide (XMPA-LCMS)

OVD HPLC12/HPLC/UV

Acetate
Formate

OVD RSK01/GC-FID

Acetylene
Propane

OVD K9305/ISE

Fluoroborate

SOLID AND HAZARDOUS CHEMICALS

OVD MSS0I/GC-MS

1-Methylnaphthalene
Benzaldehyde
Biphenyl
Caprolactam
Pentachloroethane

NELAP Accreditation by Laboratory SOP

SOLID AND HAZARDOUS CHEMICALS

OVD MSV0I/GC-MS

1.3-Butadiene
Cyclohexane
Cyclohexanone
Dimethyl disulfide
Dimethylsulfide
Ethyl-t-butylether (ETBE)
Isoprene
Methylacetate
Methylcyclohexane
n-Hexane
T-amylmethylether (TAME)

Microbac

Laboratory Report Number: L16060823

Kayla Teague
AECOM Technical Services, Inc.
16000 Dallas Parkway
Dallas, TX 75248

Please find enclosed the analytical results for the samples you submitted to Microbac Laboratories. Review and compilation of your report was completed by Microbac's Ohio Valley Division (OVD). If you have any questions, comments, or require further assistance regarding this report, please contact your service representative listed below.

Laboratory Contact:
Michelle Taylor – Client Services Specialist
(740) 373-4071
Michelle.Taylor@microbac.com

I certify that all test results meet all of the requirements of the DoD QSM and other applicable contract terms and conditions. Any exceptions are attached to this cover page or addressed in the method narratives presented in the report. All results for soil samples are reported on a 'dry-weight' basis unless specified otherwise. Analytical results for water and wastes are reported on a 'as received' basis unless specified otherwise. A statement of uncertainty for each analysis is available upon request. This laboratory report shall not be reproduced, except in full, without the written approval of Microbac Laboratories, DoD ELAP certification number 2936.01. The reported results are related only to the samples analyzed as received.

This report was certified on July 08 2016

David Vandenberg – Managing Director

State of Origin: TX
Accrediting Authority: Texas Commission on Environmental Quality ID:T104704252-07-TX
QAPP: DOD Ver 4.1

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Record of Sample Receipt and Inspection

Comments/Discrepancies

This is the record of the shipment conditions and the inspection records for the samples received and reported as a sample delivery group (SDG). All of the samples were inspected and observed to conform to our receipt policies, except as noted below.

There were no discrepancies.

Discrepancy	Resolution

Coolers

Cooler #	Temperature Gun	Temperature	COC #	Airbill #	Temp Required?
0018103	H	3.0		J2317165009	X

Inspection Checklist

#	Question	Result
1	Were shipping coolers sealed?	Yes
2	Were custody seals intact?	Yes
3	Were cooler temperatures in range of 0-6?	Yes
4	Was ice present?	Yes
5	Were COC's received/information complete/signed and dated?	Yes
6	Were sample containers intact and match COC?	Yes
7	Were sample labels intact and match COC?	Yes
8	Were the correct containers and volumes received?	Yes
9	Were samples received within EPA hold times?	Yes
10	All samples were checked for pH and met the standard. Exceptions are noted above under discrepancy. (water only)	Yes
11	Were pH ranges acceptable? (voa's excluded)	Yes
12	Were VOA samples free of headspace (less than 6mm)?	Yes

Microbac Laboratories ● Ohio Valley Division
158 Starlite Drive, Marietta, OH 45750 ● T: (740)373-4071 F: (740)373-4835
www.microbac.com

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Samples Received

Client ID	Laboratory ID	Date Collected	Date Received
126-061416	L16060823-01	06/14/2016 11:10	06/15/2016 09:56
126F-061416	L16060823-02	06/14/2016 11:10	06/15/2016 09:56
C08-061416	L16060823-03	06/14/2016 13:10	06/15/2016 09:56
C03-061416	L16060823-04	06/14/2016 14:15	06/15/2016 09:56
TRIP BLANK	L16060823-05	06/14/2016 00:01	06/15/2016 09:56

Microbac REPORT L16060823
PREPARED FOR AECOM Technical Services, Inc.
WORK ID:

1.0 Summary Data 5
1.1 Narratives 6
1.2 Certificate of Analysis 24
2.0 Full Sample Data Package 60
2.1 Volatiles Data 61
2.1.1 Volatiles GCMS Data (8260) 62
2.1.1.1 Summary Data 63
2.1.1.2 QC Summary Data 75
2.1.1.3 Sample Data 106
2.1.1.4 Standards Data 121
2.1.1.5 Raw QC Data 197
2.2 Semivolatiles Data 270
2.2.1 Semivolatiles GC/MS Data (Dioxane) 210
2.2.1.1 Summary Data 211
2.2.1.2 QC Summary Data 216
2.2.1.3 Sample Data 236
2.2.1.4 Standards Data 243
2.2.1.5 Raw QC Data 259
2.2 General Chromatography Data 270
2.2.2 6850 LC/MS Data 271
2.2.2.1 Summary Data 272
2.2.2.2 QC Summary Data 276
2.2.2.3 Sample Data 346
2.2.2.4 Standards Data 353
2.2.2.5 Raw QC Data 414
2.3 Metals Data 421
2.3.1 Metals I C P Data 422
2.3.1.1 Summary Data 423
2.3.1.2 QC Summary Data 435
2.3.1.3 Raw Data 507
2.3.2 Metals ICP-MS Data 1120
2.3.2.1 Summary Data 1121
2.3.2.2 QC Summary Data 1129
2.3.2.3 Raw Data 1187
2.3.3 Metals CVAA Data (Mercury) 2214
2.3.3.1 Summary Data 2215
2.3.3.2 QC Summary 2220
2.3.3.3 Raw Data 2242
3.0 Attachments 2247

1.0 Summary Data

1.1 Narratives

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	8260
Prep Batch Number(s):	573038	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-07 00:00:00		

Laboratory Data Package Cover Page

X	R1	Field chain-of-custody documentation;
X	R2	Sample identification cross-reference;
X	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
X	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
X	R5	Test reports/summary forms for blank samples;
X	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
X	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
X	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
X	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
X	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Franci Bolden		Analyst I	2016-07-07 20:45:33

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	8260
Prep Batch Number(s):	573038	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-07 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports					
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?	X				
Were % moisture (or solids) reported for all soil and sediment samples?	X				
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?	X				
Were surrogate percent recoveries in all samples within the laboratory QC limits?		X			1
Test reports/summary forms for blank samples	X				
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):					
Were all COCs included in the LCS?	X				
Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	8260
Prep Batch Number(s):	573038	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-07 00:00:00		

Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?	X				
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?	X				
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				
Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?	X				
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	8260
Prep Batch Number(s):	573038	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-07 00:00:00		

Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?		X			2
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?			X		
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?	X				
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?	X				
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?			X		
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?			X		
Method detection limit (MDL) studies					
Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	8260
Prep Batch Number(s):	573038	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-07 00:00:00		

Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;

2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);

3. NA = Not applicable;

4. NR = Not reviewed;

5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

1) 1,2-Dichloroethane-d4 marginally exceeded the UCL in samples 01 and 04, target compounds were not detected above the reporting limit. Reanalysis not required.

2) Bromomethane was below the LCL and 2,2-Dichloropropane exceeded the UCL in CCV WG573037-02.

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG572806	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-23 00:00:00		

Laboratory Data Package Cover Page

X	R1	Field chain-of-custody documentation;
X	R2	Sample identification cross-reference;
X	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
X	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
X	R5	Test reports/summary forms for blank samples;
X	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
X	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
X	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
X	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
X	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Eric Lawson		Chemist III	2016-06-23 18:23:03

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG572806	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-23 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports					
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?			X		
Were % moisture (or solids) reported for all soil and sediment samples?			X		
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?	X				
Were surrogate percent recoveries in all samples within the laboratory QC limits?	X				
Test reports/summary forms for blank samples	X				
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):					
Were all COCs included in the LCS?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG572806	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-23 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?	X				
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?	X				
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG572806	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-23 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?	X				
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?	X				
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?			X		
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?	X				
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?	X				
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?			X		
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?			X		
Method detection limit (MDL) studies					

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG572806	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-23 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG572806	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-23 00:00:00		

below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

There are no exceptions.

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Laboratory Data Package Cover Page

	R1	Field chain-of-custody documentation;
	R2	Sample identification cross-reference;
	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
	R5	Test reports/summary forms for blank samples;
	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Wade DeLong	Wade DeLong	Chemist I	2016-06-28 17:39:56

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports					
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?			X		
Were % moisture (or solids) reported for all soil and sediment samples?			X		
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?			X		
Were surrogate percent recoveries in all samples within the laboratory QC limits?			X		
Test reports/summary forms for blank samples	X				
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):					
Were all COCs included in the LCS?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?	X				
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?			X		
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?	X				
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?	X				
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?	X				
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?			X		
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?	X				
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?	X				
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?			X		
Method detection limit (MDL) studies					

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060823
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

1.2 Certificate of Analysis

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-01	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	126-061416	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573038	**Analyst:**	FJB	**Run Date:**	06/16/2016 20:50
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	8M413045
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	Q	10.0	5.00	2.50
Benzene	71-43-2	0.250	Q	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	Q	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	Q	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	Q	1.00	0.500	0.250
Bromoform	75-25-2	1.00	Q	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	Q	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	Q	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	Q	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	Q	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	Q	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	Q	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	Q	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	Q	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	Q	2.00	1.00	0.500
Chloroform	67-66-3	0.250	Q	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	Q	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	Q	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	Q	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	Q	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	Q	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	Q	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	Q	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	Q	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	Q	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	Q	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	Q	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	Q	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	Q	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	Q	1.00	0.500	0.250
trans-1,2-Dichloroethene	156-60-5	0.500	Q	1.00	0.500	0.250

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dichloropropane	78-87-5	0.400	Q	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	Q	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	Q	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	Q	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	Q	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	Q	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	Q	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	Q	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	1.28	J	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	Q	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	Q	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	Q	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	Q	1.00	0.250	0.125
Styrene	100-42-5	0.250	Q	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	Q	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	Q	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	Q	1.00	0.500	0.250
Toluene	108-88-3	0.500	Q	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	Q	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	Q	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	Q	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	Q	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	Q	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	Q	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	Q	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	Q	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	Q	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	Q	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	Q	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	Q	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	114	85	115	
1,2-Dichloroethane-d4	121	70	120	*
Toluene-d8	101	85	120	
4-Bromofluorobenzene	99.5	75	120	

J	The reported result is an estimated value.
Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-01	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	126-061416	**Prep Method:**	3520C	**Prep Date:**	06/15/2016 18:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573365	**Analyst:**	MES	**Run Date:**	06/21/2016 18:40
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	15M18480
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	73.6	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-01	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	126-061416	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/24/2016 21:44
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	1LM.LM35468
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	0.140	J	0.400	0.200	0.100

J	Estimated value ; the analyte concentration was less than the LOQ.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	126F-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 20:33
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	T3.062316.203318
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	0.100	U	0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Iron, Total	7439-89-6	5.93		0.200	0.100	0.0500
Potassium, Total	7440-09-7	4.36		2.00	1.00	0.500

J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	126F-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/25/2016 12:36
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/25/2016 14:03
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	T3.062516.140359
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	126F-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 21:32
Collect Date:	06/14/2016 11:10	**Dilution:**	20	**File ID:**	T3.062316.213215
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Calcium, Total	7440-70-2	296		10.0	5.00	2.50
Magnesium, Total	7439-95-4	235		20.0	10.0	5.00
Sodium, Total	7440-23-5	829		20.0	10.0	5.00

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	126F-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/28/2016 13:15
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/28/2016 20:48
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	NI.062816.204859
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.0147		0.00200	0.00100	0.000500
Cadmium, Total	7440-43-9	0.000600	U	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00231	J	0.00400	0.00200	0.00100
Cobalt, Total	7440-48-4	0.0656		0.00200	0.00100	0.000500
Copper, Total	7440-50-8	0.00314	J	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00100	U	0.00200	0.00100	0.000500
Nickel, Total	7440-02-0	0.0313		0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.000584	J	0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0168	J	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	126F-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/29/2016 12:28
Collect Date:	06/14/2016 11:10	**Dilution:**	50	**File ID:**	NI.062916.122812
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Barium, Total	7440-39-3	9.33		0.300	0.150	0.0750
Manganese, Total	7439-96-5	1.07		0.200	0.100	0.0500

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	126F-061416	**Prep Method:**	7470A	**Prep Date:**	06/16/2016 07:27
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/26/2016 12:23
Workgroup #:	WG573236	**Analyst:**	AMA	**Run Date:**	06/26/2016 13:19
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	M7.062616.131930
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	C08-061416	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573038	**Analyst:**	FJB	**Run Date:**	06/16/2016 21:19
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	8M413046
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	Q	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	5.61		1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	112	85	115	
1,2-Dichloroethane-d4	119	70	120	
Toluene-d8	104	85	120	
4-Bromofluorobenzene	100	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	C08-061416	**Prep Method:**	3520C	**Prep Date:**	06/15/2016 18:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573365	**Analyst:**	MES	**Run Date:**	06/21/2016 19:02
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	15M18481
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	44.3	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	C08-061416	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/24/2016 22:03
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	1LM.LM35469
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	0.200	U	0.400	0.200	0.100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C08-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/25/2016 12:36
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/25/2016 14:07
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	T3.062516.140759
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C08-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 20:37
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	T3.062316.203719
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	0.114	J	0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Iron, Total	7439-89-6	0.627		0.200	0.100	0.0500
Potassium, Total	7440-09-7	2.18		2.00	1.00	0.500

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C08-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 21:43
Collect Date:	06/14/2016 13:10	**Dilution:**	20	**File ID:**	T3.062316.214348
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Calcium, Total	7440-70-2	239		10.0	5.00	2.50
Magnesium, Total	7439-95-4	143		20.0	10.0	5.00
Sodium, Total	7440-23-5	761		20.0	10.0	5.00

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	C08-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/28/2016 13:15
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/28/2016 20:52
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	NI.062816.205210
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00829		0.00200	0.00100	0.000500
Cadmium, Total	7440-43-9	0.000385	J	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00247	J	0.00400	0.00200	0.00100
Cobalt, Total	7440-48-4	0.00683		0.00200	0.00100	0.000500
Copper, Total	7440-50-8	0.00295	J	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00100	U	0.00200	0.00100	0.000500
Nickel, Total	7440-02-0	0.0138		0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.00101	J	0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0250	U	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	C08-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/29/2016 12:31
Collect Date:	06/14/2016 13:10	**Dilution:**	50	**File ID:**	NI.062916.123124
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Barium, Total	7440-39-3	5.12		0.300	0.150	0.0750
Manganese, Total	7439-96-5	3.15		0.200	0.100	0.0500

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	C08-061416	**Prep Method:**	7470A	**Prep Date:**	06/16/2016 07:27
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/26/2016 12:23
Workgroup #:	WG573236	**Analyst:**	AMA	**Run Date:**	06/26/2016 13:22
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	M7.062616.132202
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	C03-061416	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573038	**Analyst:**	FJB	**Run Date:**	06/16/2016 21:48
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	8M413047
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	Q	10.0	5.00	2.50
Benzene	71-43-2	0.250	Q	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	Q	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	Q	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	Q	1.00	0.500	0.250
Bromoform	75-25-2	1.00	Q	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	Q	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	Q	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	Q	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	Q	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	Q	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	Q	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	Q	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	Q	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	Q	2.00	1.00	0.500
Chloroform	67-66-3	0.250	Q	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	Q	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	Q	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	Q	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	Q	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	Q	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	Q	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	Q	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	Q	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	Q	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	Q	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	Q	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	Q	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	Q	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	Q	1.00	0.500	0.250

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	Q	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	Q	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	Q	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	Q	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	Q	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	Q	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	Q	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	Q	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	Q	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	Q	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	Q	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	Q	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	Q	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	Q	1.00	0.250	0.125
Styrene	100-42-5	0.250	Q	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	Q	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	Q	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	Q	1.00	0.500	0.250
Toluene	108-88-3	0.500	Q	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	Q	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	Q	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	Q	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	Q	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	Q	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	Q	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	Q	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	Q	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	Q	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	Q	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	Q	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	Q	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	113	85	115	
1,2-Dichloroethane-d4	121	70	120	*
Toluene-d8	104	85	120	
4-Bromofluorobenzene	100	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	C03-061416	**Prep Method:**	3520C	**Prep Date:**	06/15/2016 18:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573365	**Analyst:**	MES	**Run Date:**	06/21/2016 19:25
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	15M18482
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	43.2	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	C03-061416	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/24/2016 22:22
Collect Date:	06/14/2016 14:15	**Dilution:**	10	**File ID:**	1LM.LM35470
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	5.48		4.00	2.00	1.00

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C03-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/25/2016 12:36
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/25/2016 14:12
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	T3.062516.141200
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C03-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 20:41
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	T3.062316.204123
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	0.100	U	0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Potassium, Total	7440-09-7	4.01		2.00	1.00	0.500

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C03-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 21:47
Collect Date:	06/14/2016 14:15	**Dilution:**	20	**File ID:**	T3.062316.214744
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Calcium, Total	7440-70-2	55.4		10.0	5.00	2.50
Iron, Total	7439-89-6	101		4.00	2.00	1.00
Magnesium, Total	7439-95-4	35.8		20.0	10.0	5.00
Sodium, Total	7440-23-5	169		20.0	10.0	5.00

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	C03-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/28/2016 13:15
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/28/2016 20:55
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	NI.062816.205522
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00684		0.00200	0.00100	0.000500
Cadmium, Total	7440-43-9	0.000600	U	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00200	U	0.00400	0.00200	0.00100
Cobalt, Total	7440-48-4	0.000578	J	0.00200	0.00100	0.000500
Copper, Total	7440-50-8	0.00200	U	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00100	U	0.00200	0.00100	0.000500
Nickel, Total	7440-02-0	0.00400	U	0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.00100	U	0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0153	J	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	C03-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/29/2016 12:34
Collect Date:	06/14/2016 14:15	**Dilution:**	50	**File ID:**	NI.062916.123435
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Barium, Total	7440-39-3	1.56		0.300	0.150	0.0750
Manganese, Total	7439-96-5	1.85		0.200	0.100	0.0500

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	C03-061416	**Prep Method:**	7470A	**Prep Date:**	06/16/2016 07:27
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/26/2016 12:23
Workgroup #:	WG573236	**Analyst:**	AMA	**Run Date:**	06/26/2016 13:24
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	M7.062616.132434
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-05	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	TRIP BLANK	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573038	**Analyst:**	FJB	**Run Date:**	06/16/2016 19:23
Collect Date:	06/14/2016 00:01	**Dilution:**	1	**File ID:**	8M413042
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	Q	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	110	85	115	
1,2-Dichloroethane-d4	114	70	120	
Toluene-d8	106	85	120	
4-Bromofluorobenzene	102	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

2.0 Full Sample Data Package

2.1 Volatiles Data

2.1.1 Volatiles GCMS Data (8260)

2.1.1.1 Summary Data

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-01	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	126-061416	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573038	**Analyst:**	FJB	**Run Date:**	06/16/2016 20:50
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	8M413045
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	Q	10.0	5.00	2.50
Benzene	71-43-2	0.250	Q	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	Q	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	Q	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	Q	1.00	0.500	0.250
Bromoform	75-25-2	1.00	Q	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	Q	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	Q	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	Q	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	Q	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	Q	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	Q	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	Q	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	Q	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	Q	2.00	1.00	0.500
Chloroform	67-66-3	0.250	Q	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	Q	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	Q	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	Q	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	Q	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	Q	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	Q	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	Q	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	Q	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	Q	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	Q	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	Q	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	Q	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	Q	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	Q	1.00	0.500	0.250
trans-1,2-Dichloroethene	156-60-5	0.500	Q	1.00	0.500	0.250

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dichloropropane	78-87-5	0.400	Q	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	Q	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	Q	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	Q	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	Q	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	Q	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	Q	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	Q	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	1.28	J	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	Q	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	Q	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	Q	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	Q	1.00	0.250	0.125
Styrene	100-42-5	0.250	Q	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	Q	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	Q	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	Q	1.00	0.500	0.250
Toluene	108-88-3	0.500	Q	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	Q	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	Q	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	Q	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	Q	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	Q	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	Q	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	Q	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	Q	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	Q	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	Q	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	Q	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	Q	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	114	85	115	
1,2-Dichloroethane-d4	121	70	120	*
Toluene-d8	101	85	120	
4-Bromofluorobenzene	99.5	75	120	

J	The reported result is an estimated value.
Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	C08-061416	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573038	**Analyst:**	FJB	**Run Date:**	06/16/2016 21:19
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	8M413046
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	Q	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	5.61		1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	112	85	115	
1,2-Dichloroethane-d4	119	70	120	
Toluene-d8	104	85	120	
4-Bromofluorobenzene	100	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	C03-061416	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573038	**Analyst:**	FJB	**Run Date:**	06/16/2016 21:48
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	8M413047
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	Q	10.0	5.00	2.50
Benzene	71-43-2	0.250	Q	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	Q	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	Q	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	Q	1.00	0.500	0.250
Bromoform	75-25-2	1.00	Q	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	Q	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	Q	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	Q	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	Q	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	Q	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	Q	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	Q	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	Q	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	Q	2.00	1.00	0.500
Chloroform	67-66-3	0.250	Q	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	Q	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	Q	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	Q	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	Q	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	Q	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	Q	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	Q	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	Q	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	Q	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	Q	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	Q	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	Q	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	Q	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	Q	1.00	0.500	0.250

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	Q	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	Q	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	Q	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	Q	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	Q	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	Q	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	Q	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	Q	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	Q	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	Q	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	Q	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	Q	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	Q	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	Q	1.00	0.250	0.125
Styrene	100-42-5	0.250	Q	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	Q	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	Q	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	Q	1.00	0.500	0.250
Toluene	108-88-3	0.500	Q	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	Q	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	Q	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	Q	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	Q	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	Q	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	Q	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	Q	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	Q	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	Q	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	Q	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	Q	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	Q	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	113	85	115	
1,2-Dichloroethane-d4	121	70	120	*
Toluene-d8	104	85	120	
4-Bromofluorobenzene	100	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-05	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	TRIP BLANK	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573038	**Analyst:**	FJB	**Run Date:**	06/16/2016 19:23
Collect Date:	06/14/2016 00:01	**Dilution:**	1	**File ID:**	8M413042
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	Q	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	110	85	115	
1,2-Dichloroethane-d4	114	70	120	
Toluene-d8	106	85	120	
4-Bromofluorobenzene	102	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.1.1.2 QC Summary Data

Example 8260 Calculations

1.0 Calculating the Response Factor (RF) from the initial calibration (ICAL) data:

RF = [(Ax) (Cis)] / [(Ais) (Cx)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured:	3399156
Cis = Concentration of the specific internal standard (ug/mL)	25
Ais = Area of the characteristic ion of the specific internal standard	846471
Cx = Concentration of the compound in the standard being measured (ug/mL)	100
RF = Calculated Response Factor	**1.0039**

2.0 Calculating the concentration (C) of a compound in water using the average RF: *

Cx = [(Ax) (Cis) (Vn)(D)] / [(Ais) (RF) (Vs)]

where:	**Example**
Ax = Area of the characteristic ion for the compound being measured	**3122498**
Cis = Concentration of the specific internal standard (ug/L)	**25**
D = Dilution factor for sample as a multiplier (10x = 10)	**1**
Ais = Area of the characteristic ion of the specific internal standard	**611048**
RF = Average RF from the ICAL	**1.004**
Vs = Purge volume of sample (mL)	**10**
Vn = Nominal purge volume of sample (mL) (10.0 mL)	**10**
Cx = Concentration of the compound in the sample being measured (ug/L)	**127.2428**

3.0 Calculating the concentration (C) of a compound in soil using the average RF: *

Cx = [(Ax) (Cis) (Wn)(D)] / [(Ais) (RF) (Ws)]

where:	**Example**
Ax = Area of the characteristic ion for the compound being measured	**3122498**
Cis = Concentration of the specific internal standard (ug/L)	**25**
D = Dilution factor for sample as a multiplier (10x = 10)	**1**
Ais = Area of the characteristic ion of the specific internal standard	**611048**
RF = Average RF from the ICAL	**1.004**
Ws = Weight of sample purged (g)	**5**
Wn = Nominal purge weight (g) (5.0 g)	**5**
Cx = Concentration of the compound in the sample being measured (ug/L)	**127.2428**

Dry weight correction:

Percent solids (PCT_S)	**50**
Cd = (Cx) (100)/PCT_S	**254.4856**

* Concentrations appearing on the instrument quantitation reports are on-column results and do not take into account initial volume, final volume, and the dilution factor.

4.0 Concentration from Linear Regression

Step 1: Retrieve Curve Data From Plot, y = mx + b

y = response ratio = response of analyte / response of IS = Ax/Ais

x = amount ratio = concentration analyte/concentration internal standard = Cx / Cis

m = slope from curve = 0.213

b = intercept from curve = - 0.00642

Step 2: Calculate y from Quantitation Report

y = 86550/593147 = 0.1459

Step 3: Solve for x

x = (y - b)/m = [(0.1459 -(- 0.00642)]/ 0.213 = 0.7152

Step 4: Solve for analyte concentration Cx

Cx = Cis (x) = (25.0)(0.7152) = 17.88

Example Spreadsheet Calculation:

Slope from curve, m:	**0.213**
Intercept from curve, b:	**-0.00642**
Area of analyte, Ax:	**86550**
Area of Internal Standard , Ais:	**593147**
Concentration of IS, Cis	**25.00**
Response Ratio:	**0.145917**
Amount Ratio:	**0.715195**
Concentration:	**17.87988**
Units of Internal Standard:	**ug/L**

5.0 Concentration from Quadratic Regression

Step 1 - Retrieve Curve Data from Plot, y = Ax^2 + Bx + C
Where:
Ax^2 + Bx + (C - y) = 0
A, B, C = constants from the ICAL quadratic regression
y = Response ratio = Area of analyte/Area of internal standard (IS)
x = Amount ratio = Concentration of analyte/concentration of IS

Step 2: Calculate y from Quantitation Report

y = Ax/Ais

Step 3: Solve for x using the quadratic formula
Ax^2 + Bx + C - y = 0

$$x = \frac{b \pm \sqrt{\left(b^2 - 4a(c - y)\right)}}{2a}$$ (Two possible solutions)

Step 4: Solve for analyte concentration Cx

Cx = (Cis)(Amount ratio)

Example Spreadsheet Calculation:

Value of A from plot:	**-0.00629**	
Value of B from plot:	**0.511**	
Value of C from plot:	**-0.0276**	
Area of unknown from quantitation report:	**293821**	
Area of IS from quantiation report:	**784848**	
Response ratio, y:	**0.374367**	
C - y:	**-0.40197**	
Root 1 - Computed amount ratio , X1:	**80.44567**	
Root 2 - Computed amount ratio , X2:	**0.794396**	use this solution
Concentration of IS, Cis:	**25.00**	
Concentration of analyte, Cx:	**19.86**	ug/L

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 051316
Analyst1: TMB Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 23
Method: 624 SOP: MSV10 Rev: 14
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID: 53752

Internal Standard: STD75899 Surrogate Standard: STD75899
CCV: STD76127 LCS: STD76109 MS/MSD: NA

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG568561

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
8M412209	WG568561-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 08:38
8M412210	WG568561-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 08:57
8M412211	WG568561-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 09:29
8M412212	WG568561-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 09:56
8M412213	WG568561-02 50ug/L CCV STD 8260	NA	1	1	STD76127	05/13/16 10:19
8M412214	WG568561-02 50ug/L CCV STD 8260	NA	1	1	STD76127	05/13/16 10:51
8M412215	WG568565-01 100ug/L A9 CCV STD 8260	NA	1	1	STD76072	05/13/16 11:36
8M412216	BFB CHECK	NA	1	1	STD76034	05/13/16 12:04
8M412217	BFB CHECK	NA	1	1	STD76034	05/13/16 12:18
8M412218	WG568561-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 12:36
8M412219	RINSE	NA	1	1		05/13/16 13:00
8M412220	WG568561-02 0.3ug/L STD 8260	NA	1	1	STD76127	05/13/16 13:29
8M412221	WG568561-03 0.4ug/L STD 8260	NA	1	1	STD76127	05/13/16 13:58
8M412222	WG568561-04 1ug/L STD 8260	NA	1	1	STD76127	05/13/16 14:27
8M412223	WG568561-05 2ug/L STD 8260	NA	1	1	STD76127	05/13/16 14:56
8M412224	WG568561-06 5ug/L STD 8260	NA	1	1	STD76127	05/13/16 15:25
8M412225	WG568561-07 20ug/L STD 8260	NA	1	1	STD76127	05/13/16 15:54
8M412226	WG568561-08 50ug/L STD 8260	NA	1	1	STD76127	05/13/16 16:23
8M412227	WG568561-09 100ug/L STD 8260	NA	1	1	STD76127	05/13/16 16:52
8M412228	WG568561-10 200ug/L STD 8260	NA	1	1	STD76127	05/13/16 17:21
8M412229	WG568561-11 300ug/L STD 8260	NA	1	1	STD76127	05/13/16 17:50
8M412230	RINSE	NA	1	1		05/13/16 18:19
8M412231	RINSE	NA	1	1		05/13/16 18:49
8M412232	WG568561-12 50ug/L ALT SRC STD 8260	NA	1	1	STD76109	05/13/16 19:17
8M412233	RINSE	NA	1	1		05/13/16 19:46

Comments

Seq.	Rerun	Dil.	Reason	Analytes
5	X			
File ID: 8M412213				
	Several compounds were low, DNR.			
6				
File ID: 8M412214				

Approved: May 20, 2016

Page: 1

Microbac Laboratories Inc.

Instrument Run Log

Instrument:	HPMS8	Dataset:	051316		
Analyst1:	TMB	Analyst2:	NA		
Method:	8260B	SOP:	MSV01	Rev:	23
Method:	624	SOP:	MSV10	Rev:	14
Method:	5030B/5030C/5035A	SOP:	PAT01	Rev:	18
Maintenance Log ID:	53752				

Internal Standard: STD75899 Surrogate Standard: STD75899

CCV: STD76127 LCS: STD76109 MS/MSD: NA

Column 1 ID: RTX502.2 Column 2 ID: NA

Workgroups: WG568561

Comments:

Comments

Seq.	Rerun	Dil.	Reason	Analytes
			Two C compounds failed. DNR. Run a curve.	
7 File ID: 8M412215				
			DNR.	

Approved: May 20, 2016

Page: 2

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 051716
Analyst1: TMB Analyst2: FJB
Method: 8260B SOP: MSV01 Rev: 23
Method: 624 SOP: MSV10 Rev: 14
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID:

Internal Standard: STD75899 Surrogate Standard: STD75899
CCV: STD76072; STD76127 LCS: STD76109; STD75801 MS/MSD: NA

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG569079; WG569083

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
8M412319	WG569084-01 50ng BFB STD 8260	NA	1	1	STD76034	05/17/16 08:53
8M412320	WG569084-02 50ug/L CCV STD 8260	NA	1	1	STD76127	05/17/16 09:18
8M412321	WG569079-01 100ug/L A9 CCV STD 8260	NA	1	1	STD76072	05/17/16 09:47
8M412322	WG569079-03 50ug/L A9/FOO STD 8260	NA	1	1	STD76072	05/17/16 10:16
8M412323	WG569079-04 20ug/L A9/FOO STD 8260	NA	1	1	STD76072	05/17/16 10:44
8M412324	WG569079-05 5ug/L A9/FOO STD 8260	NA	1	1	STD76072	05/17/16 11:13
8M412325	WG569079-06 200ug/L A9/FOO STD 8260	NA	1	1	STD76072	05/17/16 11:42
8M412326	WG569079-07 300ug/L A9/FOO STD 8260	NA	1	1	STD76072	05/17/16 12:40
8M412327	WG569079-08 400ug/L A9/FOO STD 8260	NA	1	1	STD76072	05/17/16 13:10
8M412328	WG569079-09 500ug/L A9/FOO STD 8260	NA	1	1	STD76072	05/17/16 13:39
8M412329	RINSE	NA	1	1		05/17/16 14:08
8M412330	RINSE	NA	1	1		05/17/16 14:37
8M412331	WG569079-10 100ug/L ALT SRC STD 826	NA	1	1	STD75801	05/17/16 15:06
8M412332	WG569259-01 BFB 50ng 8260	NA	1	1	STD76034	05/17/16 18:27
8M412333	WG569259-02 50ug/L CCV 8260	NA	1	1	STD76034	05/17/16 18:53
8M412334	WG569260-01 100ug/L CCV A9	NA	1	1	STD76072	05/17/16 19:22
8M412335	WG569083-01 VBLK0517 BLANK 8260	NA	1	1		05/17/16 19:51
8M412336	WG569083-02 20ug/L LCS 8260	NA	1	1	STD76109	05/17/16 20:20
8M412337	WG569083-03 20ug/L LCS2 8260	NA	1	1	STD76109	05/17/16 20:49
8M412338	WG569083-04 100ug/L LCS A9	NA	1	1	STD75801	05/17/16 21:18
8M412339	WG569083-05 100ug/L LCS2 A9	NA	1	1	STD75801	05/17/16 21:46
8M412340	L16050844-01 TB A 826-A9-SPE	5	1	1		05/17/16 22:16
8M412341	L16050844-07 EB A 826-A9-SPE	5	1	1		05/17/16 22:45
8M412342	L16050844-11 AB A 826-A9-SPE	5	1	1		05/17/16 23:14
8M412343	L16050844-02 A 826-A9-SPE	6	1	1		05/17/16 23:43
8M412344	L16050844-05 A 826-A9-SPE	6	1	1		05/18/16 00:12
8M412345	L16050844-06 A 826-A9-SPE	6	1	1		05/18/16 00:41
8M412346	L16050844-12 A 826-A9-SPE	6	1	1		05/18/16 01:10
8M412347	L16050840-01 A 826-A9-SPE	7	1	1		05/18/16 01:38
8M412348	L16050844-08 A 826-A9-SPE	7	1	1		05/18/16 02:07
8M412349	L16050844-13 A 826-A9-SPE	7	1	1		05/18/16 02:36
8M412350	L16050844-15 A 826-A9-SPE	7	1	1		05/18/16 03:05
8M412351	L16050839-01 A 826-A9-SPE	7	1	1		05/18/16 03:33
8M412352	L16050844-03 2.5X A 826-A9-SPE	7	1	2.5		05/18/16 04:02

Approved: May 20, 2016

Sarah Vandenberg

Microbac

Page: 1

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 051716
Analyst1: TMB Analyst2: FJB
Method: 8260B SOP: MSV01 Rev: 23
Method: 624 SOP: MSV10 Rev: 14
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID:

Internal Standard: STD75899 Surrogate Standard: STD75899
CCV: STD76072; STD76127 LCS: STD76109; STD75801 MS/MSD: NA

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG569079; WG569083

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
8M412353	L16050844-04 5X A 826-A9-SPE	7	1	5		05/18/16 04:31
8M412354	L16050844-09 5X A 826-A9-SPE	7	1	5		05/18/16 05:01
8M412355	L16050844-10 5X A 826-A9-SPE	7	1	5		05/18/16 05:30
8M412356	L16050844-14 50X A 826-A9-SPE	7	1	50		05/18/16 05:59
8M412357	RINSE	NA	1	1		05/18/16 06:28

Comments

Seq.	Rerun	Dil.	Reason	Analytes
2				
File ID: 8M412320				
	DNR. Running an a9/foo curve.			
32	X	10	Over Calibration Range	CIS12-DCE
File ID: 8M412350				
40	X	2.5	Analyzed too dilute	
File ID: 8M412353				
	DNR.			
37	X	25	Over Calibration Range	TCE
File ID: 8M412355				
38	X	20	Analyzed too dilute	
File ID: 8M412356				
	DNR.			

Approved: May 20, 2016

Sarah Vandenberg

Page: 2

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 061616
Analyst1: ADC Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 23
Method: OVAP SOP: MSV01/PAT01 Rev: 0
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID:

Internal Standard: STD75370 Surrogate Standard: STD76370
CCV: STD76578 LCS: STD76597 MS/MSD: STD76597

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573038

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
8M413031	WG573037-01 BFB 50ng 8260	NA	1	1	STD76496	06/16/16 14:02
8M413032	WG573037-02 50ug/L CCV 8260	NA	1	1	STD76578	06/16/16 14:25
8M413033	WG573XXX-01 100ug/L CCV A9	NA	1	1	STD7XXXX	06/16/16 14:54
8M413034	WG573038-01 BLANK 6/16 8260	NA	1	1		06/16/16 15:23
8M413035	WG573038-02 20ug/L LCS 8260	NA	1	1	STD76597	06/16/16 15:52
8M413036	WG573038-03 20ug/L LCS2 8260	NA	1	1	STD76597	06/16/16 16:21
8M413037	L16060637-04 TB A 826-SPE	<2	1	1		06/16/16 16:58
8M413038	L16060756-05 TB A 826-SPE	<2	1	1		06/16/16 17:27
8M413039	L16060765-01 TB A 826-SPE	<2	1	1		06/16/16 17:56
8M413040	L16060730-04 TB A 826-LOW	<2	1	1		06/16/16 18:25
8M413041	L16060732-03 TB A 826-LOW	<2	1	1		06/16/16 18:54
8M413042	L16060823-05 TB A 826-LOW	<2	1	1		06/16/16 19:23
8M413043	L16060616-01 B 826-LOW	<2	1	1		06/16/16 19:52
8M413044	L16060732-01 A 826-LOW	<2	1	1		06/16/16 20:21
8M413045	L16060823-01 A 826-LOW	<2	1	1		06/16/16 20:50
8M413046	L16060823-03 A 826-LOW	<2	1	1		06/16/16 21:19
8M413047	L16060823-04 A 826-LOW	<2	1	1		06/16/16 21:48
8M413048	L16060730-03 A 826-LOW	<2	1	1		06/16/16 22:17
8M413049	L16060765-02 A 826-SPE	<2	1	1		06/16/16 22:46
8M413050	L16060765-03 A 826-SPE	<2	1	1		06/16/16 23:15
8M413051	L16060765-04 A 826-SPE	<2	1	1		06/16/16 23:43
8M413052	L16060730-01 A 826-SPE	<2	1	1		06/17/16 00:12
8M413053	L16060730-02 A 826-SPE	<2	1	1		06/17/16 00:41
8M413054	L16060637-03 A 826-SPE	<2	1	1		06/17/16 01:10
8M413055	L16060756-04 A 826-SPE	<2	1	1		06/17/16 01:38
8M413056	RINSE	NA	1	1		06/17/16 02:07
8M413057	RINSE	NA	1	1		06/17/16 02:37
8M413058	RINSE	NA	1	1		06/17/16 03:05

Comments

Seq.	Rerun	Dil.	Reason	Analytes
14	X		Carry-over contamination	
File ID: 8M413044				

Approved: June 23, 2016

Sarah Vandenberg

Page: 1

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 061616
Analyst1: ADC Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 23
Method: OVAP SOP: MSV01/PAT01 Rev: 0
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID:

Internal Standard: STD75370 Surrogate Standard: STD76370
CCV: STD76578 LCS: STD76597 MS/MSD: STD76597

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573038

Comments:

Comments

Seq.	Rerun	Dil.	Reason	Analytes
	L16060732-01			
15 File ID: 8M413045				
	L16060823-01 high surrogate but sample is ND.			
17 File ID: 8M413047	X		Surrogate standard failure	
	L16060823-04			
18 File ID: 8M413048	X		Surrogate standard failure	
	L16060730-03			
24 File ID: 8M413054	X	20	Over Calibration Range	CFORM, CTET, TCE
	L16060637-03			
25 File ID: 8M413055	X	20	Over Calibration Range	CB
	L16060756-04			

Approved: June 23, 2016

Sarah Vandenberg

Page: 2

Microbac Laboratories Inc.

Data Checklist

Date: 13-MAY-2016
Analyst: TMB
Analyst: NA
Method: 8260B/624
Instrument: HPMS8
Curve Workgroup: NA
Runlog ID: 75126
Analytical Workgroups: WG568561

System Performance Check	NA
BFB	X
Initial Calibration	X
Average RF	X
Linear Reg or Higher Order Curve	X
Second Source standard % Difference	X
Continuing Calibration /Check Standards	X
Project/Client Specific Requirements	X
Special Standards	X
Blanks	X
TCL's	X
Surrogates	X
LCS (Laboratory Control Sample)	X
Recoveries	X
Surrogates	X
MS/MSD/Duplicates	NA
Samples	X
TCL Hits	X
Spectra of TCL Hits	X
Surrogates	TMB
Internal Standards Criteria	X
Library Searches	NA
Calculations & Correct Factors	X
Dilutions Run	NA
Reruns	NA
Manual Integrations	NA
Case Narrative	X
Results Reporting/Data Qualifiers	X
KOBRA Workgroup Data	X
Check for Completeness	X
Primary Reviewer	TMB
Secondary Reviewer	FJB
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Check the reasonableness of the results	X

Primary Reviewer:
17-MAY-2016
Tiffany Bailey

Secondary Reviewer:
20-MAY-2016

CHECKLIST1 - Modified 03/05/2008
Generated: MAY-20-2016 11:18:36

Microbac Laboratories Inc.

Data Checklist

Date: 17-MAY-2016
Analyst: TMB
Analyst: FJB
Method: 8260B/624
Instrument: HPMS8
Curve Workgroup: NA
Runlog ID: 75174
Analytical Workgroups: WG569079; WG569083

System Performance Check	NA
BFB	X
Initial Calibration	X
Average RF	X
Linear Reg or Higher Order Curve	X
Second Source standard % Difference	X
Continuing Calibration /Check Standards	X
Project/Client Specific Requirements	X
Special Standards	X
Blanks	X
TCL's	X
Surrogates	X
LCS (Laboratory Control Sample)	X
Recoveries	X
Surrogates	X
MS/MSD/Duplicates	NA
Samples	X
TCL Hits	X
Spectra of TCL Hits	TMB
Surrogates	X
Internal Standards Criteria	X
Library Searches	NA
Calculations & Correct Factors	X
Dilutions Run	NA
Reruns	X
Manual Integrations	NA
Case Narrative	X
Results Reporting/Data Qualifiers	X
KOBRA Workgroup Data	X
Check for Completeness	X
Primary Reviewer	TMB
Secondary Reviewer	SAV
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Check the reasonableness of the results	X

Primary Reviewer:
19-MAY-2016
Tiffany Bailey

Secondary Reviewer:
20-MAY-2016
Sarah Vandenberg

CHECKLIST1 - Modified 03/05/2008
Generated: MAY-20-2016 11:05:37

Microbac Laboratories Inc.

Data Checklist

Date: 16-JUN-2016
Analyst: ADC
Analyst: NA
Method: 8260
Instrument: HPMS8
Curve Workgroup: NA
Runlog ID: 75860
Analytical Workgroups: WG573038

System Performance Check	NA
BFB	X
Initial Calibration	X
Average RF	X
Linear Reg or Higher Order Curve	X
Second Source standard % Difference	X
Continuing Calibration /Check Standards	X
Project/Client Specific Requirements	X
Special Standards	NA
Blanks	X
TCL's	X
Surrogates	X
LCS (Laboratory Control Sample)	X
Recoveries	X
Surrogates	X
MS/MSD/Duplicates	X
Samples	X
TCL Hits	X
Spectra of TCL Hits	ADC
Surrogates	X
Internal Standards Criteria	X
Library Searches	NA
Calculations & Correct Factors	X
Dilutions Run	X
Reruns	X
Manual Integrations	NA
Case Narrative	X
Results Reporting/Data Qualifiers	X
KOBRA Workgroup Data	X
Check for Completeness	X
Primary Reviewer	ADC
Secondary Reviewer	SAV
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Check the reasonableness of the results	X

Primary Reviewer:
22-JUN-2016

Secondary Reviewer:
23-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-23-2016 13:34:00

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 8260B

AAB#: WG573038

Login Number: L16060823

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
126-061416	01	06/14/16					06/16/2016	2.4	14		06/16/16	2.4	14	
C08-061416	03	06/14/16					06/16/2016	2.3	14		06/16/16	2.3	14	
C03-061416	04	06/14/16					06/16/2016	2.3	14		06/16/16	2.3	14	
TRIP BLANK	05	06/14/16					06/16/2016	2.8	14		06/16/16	2.8	14	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4819682
Report generated 07/07/2016 16:44

Microbac Laboratories Inc.
SURROGATE STANDARDS

Login Number: L16060823
Instrument Id: HPMS8
Workgroup (AAB#): WG573038

Method: 8260
CAL ID: HPMS8-13-MAY-16
Matrix: Water

Sample Number	Dilution	Tag	1	2	3	4
L16060823-01	1.00	01	121	114	99.5	101
L16060823-03	1.00	01	119	112	100	104
L16060823-04	1.00	01	121	113	100	104
L16060823-05	1.00	01	114	110	102	106
WG573038-01	1.00	01	114	109	102	106
WG573038-02	1.00	01	118	110	99.2	104
WG573038-03	1.00	01	114	108	98.9	104

Surrogates	Surrogate Limits		
1 - 1,2-Dichloroethane-d4	70	-	120
2 - Dibromofluoromethane	85	-	115
3 - 4-Bromofluorobenzene	75	-	120
4 - Toluene-d8	85	-	120

Underline = Result out of surrogate limits

DL = surrogate diluted out
ND = surrogate not detected

SURROGATES - Modified 03/06/2008
PDF File ID: 4819690
Report generated: 07/07/2016 16:44

METHOD BLANK SUMMARY

Login Number: L16060823
Blank File ID: 8M413034
Prep Date: 06/16/16 15:23
Analyzed Date: 06/16/16 15:23
Analyst: FJB

Work Group: WG573038
Blank Sample ID: WG573038-01
Instrument ID: HPMS8
Method: 8260B

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573038-02	8M413035	06/16/16 15:52	01
LCS2	WG573038-03	8M413036	06/16/16 16:21	01
TRIP BLANK	L16060823-05	8M413042	06/16/16 19:23	01
126-061416	L16060823-01	8M413045	06/16/16 20:50	01
C08-061416	L16060823-03	8M413046	06/16/16 21:19	01
C03-061416	L16060823-04	8M413047	06/16/16 21:48	01

Report Name: BLANK_SUMMARY
PDF File ID: 4819683
Report generated 07/07/2016 16:44

Microbac Laboratories Inc.

METHOD BLANK REPORT

Login Number: L16060823 Prep Date: 06/16/16 15:23 Sample ID: WG573038-01
Instrument ID: HPMS8 Run Date: 06/16/16 15:23 Prep Method: 5030B/5030C/503
File ID: 8M413034 Analyst: FJB Method: 8260B
Workgroup (AAB#): WG573038 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS8-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Acetone	2.50	10.0	2.50	1	U
Benzene	0.125	1.00	0.125	1	U
Bromobenzene	0.125	1.00	0.125	1	U
Bromochloromethane	0.200	1.00	0.200	1	U
Bromodichloromethane	0.250	1.00	0.250	1	U
Bromoform	0.500	2.00	0.500	1	U
Bromomethane	0.500	2.00	0.500	1	U
2-Butanone	2.50	10.0	2.50	1	U
n-Butylbenzene	0.250	1.00	0.250	1	U
sec-Butylbenzene	0.250	1.00	0.250	1	U
tert-Butylbenzene	0.250	1.00	0.250	1	U
Carbon disulfide	0.500	2.00	0.500	1	U
Carbon tetrachloride	0.250	1.00	0.250	1	U
Chlorobenzene	0.125	1.00	0.125	1	U
Chlorodibromomethane	0.250	1.00	0.250	1	U
Chloroethane	0.500	2.00	0.500	1	U
Chloroform	0.125	1.00	0.125	1	U
Chloromethane	0.500	2.00	0.500	1	U
2-Chlorotoluene	0.125	1.00	0.125	1	U
4-Chlorotoluene	0.250	1.00	0.250	1	U
1,2-Dibromo-3-chloropropane	1.00	5.00	1.00	1	U
1,2-Dibromoethane	0.250	1.00	0.250	1	U
Dibromomethane	0.250	1.00	0.250	1	U
1,2-Dichlorobenzene	0.125	1.00	0.125	1	U
1,3-Dichlorobenzene	0.250	1.00	0.250	1	U
1,4-Dichlorobenzene	0.125	1.00	0.125	1	U
Dichlorodifluoromethane	0.250	1.00	0.250	1	U
1,1-Dichloroethane	0.125	1.00	0.125	1	U
1,2-Dichloroethane	0.250	1.00	0.250	1	U
1,1-Dichloroethene	0.500	2.00	0.500	1	U
cis-1,2-Dichloroethene	0.250	1.00	0.250	1	U
trans-1,2-Dichloroethene	0.250	1.00	0.250	1	U
1,2-Dichloropropane	0.200	1.00	0.200	1	U
1,3-Dichloropropane	0.200	1.00	0.200	1	U
2,2-Dichloropropane	0.250	1.00	0.250	1	U
cis-1,3-Dichloropropene	0.250	1.00	0.250	1	U
trans-1,3-Dichloropropene	0.500	2.00	0.500	1	U
1,1-Dichloropropene	0.250	1.00	0.250	1	U
Ethylbenzene	0.250	1.00	0.250	1	U
2-Hexanone	2.50	10.0	2.50	1	U
Hexachlorobutadiene	0.250	1.00	0.250	1	U
Isopropylbenzene	0.250	1.00	0.250	1	U

Report Name:BLANK
PDF ID: 4819684
07-JUL-2016 16:44

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060823 Prep Date: 06/16/16 15:23 Sample ID: WG573038-01
Instrument ID: HPMS8 Run Date: 06/16/16 15:23 Prep Method: 5030B/5030C/503
File ID: 8M413034 Analyst: FJB Method: 8260B
Workgroup (AAB#): WG573038 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS8-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
p-Isopropyltoluene	0.250	1.00	0.250	1	U
4-Methyl-2-pentanone	2.50	10.0	2.50	1	U
Methylene chloride	0.250	1.00	0.250	1	U
Naphthalene	0.200	1.00	0.200	1	U
n-Propylbenzene	0.125	1.00	0.125	1	U
Styrene	0.125	1.00	0.125	1	U
1,1,1,2-Tetrachloroethane	0.250	1.00	0.250	1	U
1,1,2,2-Tetrachloroethane	0.200	1.00	0.200	1	U
Tetrachloroethene	0.250	1.00	0.250	1	U
Toluene	0.250	1.00	0.250	1	U
1,2,3-Trichlorobenzene	0.150	1.00	0.150	1	U
1,2,4-Trichlorobenzene	0.200	1.00	0.200	1	U
1,1,1-Trichloroethane	0.250	1.00	0.250	1	U
1,1,2-Trichloroethane	0.250	1.00	0.250	1	U
Trichloroethene	0.250	1.00	0.250	1	U
Trichlorofluoromethane	0.250	1.00	0.250	1	U
1,2,3-Trichloropropane	0.500	2.00	0.500	1	U
1,2,4-Trimethylbenzene	0.250	1.00	0.250	1	U
1,3,5-Trimethylbenzene	0.250	1.00	0.250	1	U
Vinyl chloride	0.250	1.00	0.250	1	U
o-Xylene	0.250	1.00	0.250	1	U
m-,p-Xylene	0.500	2.00	0.500	1	U

Surrogates	% Recovery	Surrogate Limits	Qualifier
Dibromofluoromethane	109	85 - 115	PASS
1,2-Dichloroethane-d4	114	70 - 120	PASS
Toluene-d8	106	85 - 120	PASS
4-Bromofluorobenzene	102	75 - 120	PASS

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4819684
07-JUL-2016 16:44

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060823 Analyst: FJB Prep Method: 5030B/5030C/503
Instrument ID: HPMS8 Matrix: Water Method: 8260B
Workgroup (AAB#): WG573038 Units: ug/L
QC Key: DOD4 Lot #: STD76597

Sample ID: WG573038-02 LCS File ID: 8M413035 Run Date: 06/16/2016 15:52
Sample ID: WG573038-03 LCS2 File ID: 8M413036 Run Date: 06/16/2016 16:21

Analytes	LCS			LCS2			%RPD	%Rec Limits	RPD Lmt	Q
	Known	Found	% REC	Known	Found	% REC				
1,1,1,2-Tetrachloroethane	20.0	20.8	104	20.0	20.9	105	0.576	80 - 130	30	
1,1,1-Trichloroethane	20.0	22.5	112	20.0	23.2	116	3.45	65 - 130	30	
1,1,2,2-Tetrachloroethane	20.0	17.7	88.5	20.0	18.1	90.5	2.29	65 - 130	30	
1,1,2-Trichloroethane	20.0	19.6	97.9	20.0	19.6	98.1	0.148	75 - 125	30	
1,1-Dichloroethane	20.0	19.5	97.5	20.0	19.9	99.5	2.06	70 - 135	30	
1,1-Dichloroethene	20.0	18.9	94.6	20.0	19.6	98.2	3.75	70 - 130	30	
1,1-Dichloropropene	20.0	20.6	103	20.0	21.0	105	2.03	75 - 130	30	
1,2,3-Trichlorobenzene	20.0	19.0	94.8	20.0	19.4	96.8	2.05	55 - 140	30	
1,2,3-Trichloropropane	20.0	19.8	99.1	20.0	19.6	97.8	1.36	75 - 125	30	
1,2,4-Trichlorobenzene	20.0	19.4	96.9	20.0	20.1	100	3.60	65 - 135	30	
1,2,4-Trimethylbenzene	20.0	19.9	99.5	20.0	20.3	102	2.19	75 - 130	30	
1,2-Dibromo-3-chloropropane	20.0	19.7	98.7	20.0	21.4	107	7.93	50 - 130	30	
1,2-Dibromoethane	20.0	20.1	100	20.0	19.9	99.3	1.19	80 - 120	30	
1,2-Dichlorobenzene	20.0	19.7	98.6	20.0	20.3	102	3.14	70 - 120	30	
1,2-Dichloroethane	20.0	23.3	117	20.0	23.5	118	0.816	70 - 130	30	
1,2-Dichloropropane	20.0	20.4	102	20.0	21.0	105	3.16	75 - 125	30	
1,3,5-Trimethylbenzene	20.0	20.3	102	20.0	20.9	105	3.05	75 - 130	30	
1,3-Dichlorobenzene	20.0	20.0	99.9	20.0	20.6	103	3.29	75 - 125	30	
1,3-Dichloropropane	20.0	20.7	103	20.0	21.0	105	1.29	75 - 125	30	
1,4-Dichlorobenzene	20.0	19.7	98.6	20.0	20.4	102	3.14	75 - 125	30	
2,2-Dichloropropane	20.0	23.0	115	20.0	23.5	118	2.50	70 - 135	30	
2-Butanone	20.0	20.2	101	20.0	20.8	104	3.17	30 - 150	30	
2-Chlorotoluene	20.0	20.3	102	20.0	20.6	103	1.51	75 - 125	30	
2-Hexanone	20.0	18.6	93.0	20.0	17.8	89.1	4.24	55 - 130	30	
4-Chlorotoluene	20.0	19.7	98.5	20.0	20.0	100	1.77	75 - 130	30	
4-Methyl-2-pentanone	20.0	19.8	98.8	20.0	20.7	103	4.57	60 - 135	30	
Acetone	20.0	21.8	109	20.0	24.1	121	10.2	40 - 140	30	
Benzene	20.0	20.4	102	20.0	21.1	105	3.43	80 - 120	30	
Bromobenzene	20.0	19.2	95.9	20.0	19.7	98.4	2.64	75 - 125	30	
Bromochloromethane	20.0	20.7	103	20.0	20.6	103	0.508	65 - 130	30	
Bromodichloromethane	20.0	21.6	108	20.0	22.5	112	4.14	75 - 120	30	
Bromoform	20.0	20.0	99.8	20.0	20.0	99.9	0.134	70 - 130	30	
Bromomethane	20.0	14.0	70.2	20.0	15.7	78.3	11.0	30 - 145	30	
Carbon disulfide	20.0	18.1	90.6	20.0	18.6	92.9	2.55	35 - 160	30	
Carbon tetrachloride	20.0	22.5	112	20.0	23.1	115	2.71	65 - 140	30	
Chlorobenzene	20.0	20.4	102	20.0	20.7	104	1.62	80 - 120	30	
Chloroethane	20.0	20.4	102	20.0	21.7	109	6.06	60 - 135	30	
Chloroform	20.0	21.2	106	20.0	21.7	108	2.28	65 - 135	30	
Chloromethane	20.0	18.9	94.7	20.0	20.2	101	6.46	40 - 125	30	
cis-1,2-Dichloroethene	20.0	20.5	102	20.0	21.2	106	3.58	70 - 125	30	

LCS_LCS2 - Modified 03/06/2008
PDF File ID: 4819685
Report generated: 07/07/2016 16:44

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060823 Analyst: FJB Prep Method: 5030B/5030C/503
Instrument ID: HPMS8 Matrix: Water Method: 8260B
Workgroup (AAB#): WG573038 Units: ug/L
QC Key: DOD4 Lot #: STD76597

Sample ID: WG573038-02 LCS File ID: 8M413035 Run Date: 06/16/2016 15:52
Sample ID: WG573038-03 LCS2 File ID: 8M413036 Run Date: 06/16/2016 16:21

Analytes	LCS			LCS2			%RPD	%Rec Limits	RPD Lmt	Q
	Known	Found	% REC	Known	Found	% REC				
cis-1,3-Dichloropropene	20.0	23.1	115	20.0	23.5	117	1.55	70 - 130	30	
Chlorodibromomethane	20.0	20.5	102	20.0	20.8	104	1.79	60 - 135	30	
Dibromomethane	20.0	19.9	99.3	20.0	19.9	99.6	0.358	75 - 125	30	
Dichlorodifluoromethane	20.0	20.8	104	20.0	21.8	109	4.81	30 - 155	30	
Ethylbenzene	20.0	19.9	99.3	20.0	19.9	99.5	0.215	75 - 125	30	
Hexachlorobutadiene	20.0	21.3	106	20.0	21.7	108	1.84	50 - 140	30	
Isopropylbenzene	20.0	21.6	108	20.0	22.0	110	1.89	75 - 125	30	
m-,p-Xylene	40.0	40.9	102	40.0	41.7	104	2.14	75 - 130	30	
Methylene chloride	20.0	19.3	96.4	20.0	19.9	99.5	3.13	55 - 140	30	
n-Butylbenzene	20.0	19.4	97.2	20.0	19.8	99.2	2.07	70 - 135	30	
n-Propylbenzene	20.0	20.4	102	20.0	21.0	105	2.87	70 - 130	30	
Naphthalene	20.0	17.4	87.0	20.0	18.1	90.5	3.97	55 - 140	30	
o-Xylene	20.0	21.1	106	20.0	21.4	107	1.39	80 - 120	30	
p-Isopropyltoluene	20.0	20.0	100	20.0	20.9	104	4.05	75 - 130	30	
sec-Butylbenzene	20.0	20.2	101	20.0	20.7	104	2.34	70 - 125	30	
Styrene	20.0	20.7	103	20.0	21.1	106	2.22	65 - 135	30	
tert-Butylbenzene	20.0	19.7	98.5	20.0	20.4	102	3.70	70 - 130	30	
Tetrachloroethene	20.0	20.1	101	20.0	20.5	102	1.70	45 - 150	30	
Toluene	20.0	19.7	98.6	20.0	20.1	101	1.94	75 - 120	30	
trans-1,2-Dichloroethene	20.0	20.1	100	20.0	20.4	102	1.57	60 - 140	30	
trans-1,3-Dichloropropene	20.0	20.9	104	20.0	21.1	105	0.925	55 - 140	30	
Trichloroethene	20.0	21.8	109	20.0	22.5	112	3.24	70 - 125	30	
Trichlorofluoromethane	20.0	22.3	111	20.0	22.8	114	2.49	60 - 145	30	
Vinyl chloride	20.0	23.0	115	20.0	24.1	121	4.97	50 - 145	30	

Surogates	LCS	LCS2	Surrogate Limits	Qualifier
	% Recovery	% Recovery		
1,2-Dichloroethane-d4	118	114	70 - 120	PASS
Dibromofluoromethane	110	108	85 - 115	PASS
4-Bromofluorobenzene	99.2	98.9	75 - 120	PASS
Toluene-d8	104	104	85 - 120	PASS

* EXCEEDS %REC LIMIT

\# EXCEEDS RPD LIMIT

LCS_LCS2 - Modified 03/06/2008
PDF File ID: 4819685
Report generated: 07/07/2016 16:44

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

BFB

Login Number: L16060823
Instrument: HPMS8
Analyst: TMB
Workgroup: WG568561

Tune ID: WG568561-01
Run Date: 05/13/2016
Run Time: 12:36
File ID: 8M412218
Cal ID: HPMS8 -

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50.0	95.0	15.0	40.0	16.7	6575	PASS
75.0	95.0	30.0	60.0	51.3	20171	PASS
95.0	95.0	100	100	100	39312	PASS
96.0	95.0	5.00	9.00	6.74	2649	PASS
173	174	0	2.00	0.363	122	PASS
174	95.0	50.0	100	85.5	33616	PASS
175	174	5.00	9.00	8.17	2746	PASS
176	174	95.0	101	99.2	33360	PASS
177	176	5.00	9.00	6.80	2267	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG568561-02	STD	01	05/13/2016 13:29	
WG568561-03	STD	01	05/13/2016 13:58	
WG568561-04	STD	01	05/13/2016 14:27	
WG568561-05	STD	01	05/13/2016 14:56	
WG568561-06	STD	01	05/13/2016 15:25	
WG568561-07	STD	01	05/13/2016 15:54	
WG568561-08	STD-CCV	01	05/13/2016 16:23	
WG568561-09	STD	01	05/13/2016 16:52	
WG568561-10	STD	01	05/13/2016 17:21	
WG568561-11	STD	01	05/13/2016 17:50	
WG568561-12	SSCV	01	05/13/2016 19:17	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4819687
Report generated 07/07/2016 16:44

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

BFB

Login Number: L16060823
Instrument: HPMS8
Analyst: TMB
Workgroup: WG569084

Tune ID: WG569084-01
Run Date: 05/17/2016
Run Time: 08:53
File ID: 8M412319
Cal ID: HPMS8 - 13-MAY-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50.0	95.0	15.0	40.0	19.1	6316	PASS
75.0	95.0	30.0	60.0	54.7	18130	PASS
95.0	95.0	100	100	100	33138	PASS
96.0	95.0	5.00	9.00	6.76	2241	PASS
173	174	0	2.00	0	0	PASS
174	95.0	50.0	100	72.2	23934	PASS
175	174	5.00	9.00	7.88	1886	PASS
176	174	95.0	101	96.8	23159	PASS
177	176	5.00	9.00	7.15	1655	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG569079-01	STD-CCV	01	05/17/2016 09:47	
WG569079-03	STD	01	05/17/2016 10:16	
WG569079-04	STD	01	05/17/2016 10:44	
WG569079-05	STD	01	05/17/2016 11:13	
WG569079-06	STD	01	05/17/2016 11:42	
WG569079-07	STD	01	05/17/2016 12:40	
WG569079-08	STD	01	05/17/2016 13:10	
WG569079-09	STD	01	05/17/2016 13:39	
WG569079-10	SSCV	01	05/17/2016 15:06	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4819687
Report generated 07/07/2016 16:44

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

BFB

Login Number: L16060823
Instrument: HPMS8
Analyst: FJB
Workgroup: WG573037

Tune ID: WG573037-01
Run Date: 06/16/2016
Run Time: 14:02
File ID: 8M413031
Cal ID: HPMS8 - 13-MAY-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50.0	95.0	15.0	40.0	19.9	6352	PASS
75.0	95.0	30.0	60.0	57.2	18249	PASS
95.0	95.0	100	100	100	31890	PASS
96.0	95.0	5.00	9.00	7.42	2366	PASS
173	174	0	2.00	1.50	352	PASS
174	95.0	50.0	100	73.7	23499	PASS
175	174	5.00	9.00	7.96	1870	PASS
176	174	95.0	101	97.3	22869	PASS
177	176	5.00	9.00	6.73	1538	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573037-02	CCV	01	06/16/2016 14:25	
WG573038-01	BLANK	01	06/16/2016 15:23	
WG573038-02	LCS	01	06/16/2016 15:52	
WG573038-03	LCS2	01	06/16/2016 16:21	
L16060823-05	TRIP BLANK	01	06/16/2016 19:23	
L16060823-01	126-061416	01	06/16/2016 20:50	
L16060823-03	C08-061416	01	06/16/2016 21:19	
L16060823-04	C03-061416	01	06/16/2016 21:48	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4819687
Report generated 07/07/2016 16:44

Calibration Table Report
Method: 8260WT.M
Title: Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Calibration: Sat May 14 18:08:06 2016
Curve: WG568561
Calibration Files

	Compound	0.3 8M412220.D	0.4 8M412221.D	1 8M412222.D	2 8M412223.D	5 8M412224.D	20 8M412225.D	50 8M412226.D	100 8M412227.D	200 8M412228.D	300 8M412229.D	Avg	%RSD	Linear	Quad
I	Fluorobenzene	ISTD													
T	Dichlorodifluoromethane		0.427	0.390	0.397	0.364	0.468	0.439	0.452	0.427	0.408	0.419	7.751		
P	Chloromethane			0.264	0.257	0.223	0.253	0.240	0.240	0.239	0.241	0.245	5.336		
C	Vinyl Chloride		0.274	0.296	0.297	0.277	0.299	0.292	0.299	0.292	0.288	0.290	3.191		
T	1,3-Butadiene				0.218	0.208	0.242	0.251	0.277	0.254	0.254	0.243	9.644		
T	Bromomethane			0.129	0.133	0.127	0.162	0.167	0.187	0.205	0.212	0.165	20.449		1.000
T	Chloroethane		0.173	0.157	0.153	0.143	0.164	0.157	0.163	0.164	0.164	0.160	5.354		
T	Trichlorofluoromethane		0.544	0.538	0.559	0.506	0.570	0.538	0.551	0.531	0.525	0.540	3.520		
T	Diethyl ether			0.130	0.138	0.123	0.132	0.125	0.128		0.127	0.129	3.866		
T	Isoprene		0.335	0.351	0.334	0.293	0.334	0.320	0.334	0.322	0.329	0.328	4.867		
T	Acrolein				0.012	0.010	0.010	0.011	0.012		0.014	0.012	12.987		
T	1,1,2-Trichloro-1,2,2-Trifluoroethane			0.256	0.269	0.240	0.260	0.249	0.257	0.251	0.250	0.254	3.397		
T	Acetone					0.023	0.030	0.029	0.030	0.028	0.028	0.028	9.132		
C	1,1-Dichloroethene		0.424	0.407	0.406	0.387	0.416	0.398	0.409	0.400	0.393	0.404	2.830		
T	Tert-Butyl Alcohol				0.010	0.009	0.009	0.009	0.009		0.009	0.009	2.872		
T	Dimethyl Sulfide		0.133	0.142	0.152	0.137	0.145	0.142	0.151	0.147	0.151	0.145	4.534		
T	Iodomethane				0.019	0.024	0.066	0.101	0.129	0.135	0.131	0.087	58.244		0.998
T	Methyl acetate					0.075	0.076	0.074	0.078	0.076	0.078	0.076	2.221		
T	Methylene Chloride			0.274	0.265	0.244	0.263	0.250	0.255	0.254	0.252	0.257	3.686		
T	Carbon Disulfide			0.888	0.922	0.801	0.873	0.816	0.854	0.813	0.802	0.846	5.346		
T	Acrylonitrile			0.027	0.035	0.033	0.038	0.038	0.039		0.044	0.036	14.671		
T	Methyl Tert Butyl Ether			0.534	0.543	0.521	0.586	0.579	0.585	0.560	0.546	0.557	4.455		
T	trans-1,2-Dichloroethene		0.389	0.395	0.364	0.352	0.381	0.367	0.372	0.364	0.358	0.371	3.858		
T	n-Hexane					0.265	0.280	0.270	0.286	0.275	0.279	0.276	2.732		
T	Diisopropyl ether			0.600	0.607	0.548	0.588	0.572	0.571		0.567	0.579	3.559		
T	Vinyl Acetate				0.256	0.243	0.270	0.275	0.266	0.271	0.275	0.265	4.509		
P	1,1-Dichloroethane		0.441	0.463	0.486	0.434	0.488	0.464	0.470	0.459	0.447	0.461	4.012		
T	Ethyl-Tert-Butyl ether			0.616	0.645	0.590	0.641	0.628	0.619		0.603	0.620	3.188		
T	2-Butanone					0.035	0.042	0.043	0.045	0.041	0.043	0.041	8.268		
T	Propionitrile			0.010	0.012	0.012	0.013	0.012	0.013		0.013	0.012	8.354		
T	2,2-Dichloropropane		0.471	0.462	0.520	0.476	0.503	0.494	0.497	0.483	0.470	0.486	3.839		
T	cis-1,2-Dichloroethene		0.299	0.298	0.291	0.267	0.295	0.282	0.289	0.283	0.277	0.287	3.685		
C	Chloroform	0.563	0.567	0.581	0.536	0.509	0.566	0.543	0.546	0.529	0.506	0.545	4.619		
T	1-Bromopropane			0.034	0.040	0.040	0.043	0.041	0.043	0.043	0.044	0.041	7.649		
T	Bromochloromethane		0.131	0.164	0.147	0.137	0.154	0.148	0.151	0.148	0.145	0.147	6.598		
T	Tetrahydrofuran			0.034	0.028	0.025	0.026	0.026	0.026		0.026	0.027	11.724		
S	Dibromofluoromethane			0.249	0.261	0.257	0.280	0.271	0.284	0.271	0.275	0.269	4.465		
T	1,1,1-Trichloroethane		0.492	0.537	0.515	0.496	0.545	0.525	0.533	0.515	0.499	0.517	3.660		
T	Cyclohexane			0.373	0.369	0.335	0.359	0.344	0.366	0.356	0.364	0.358	3.619		
T	1,1-Dichloropropene			0.414	0.399	0.369	0.403	0.395	0.399	0.392	0.382	0.394	3.488		
T	Tert-Amyl-Methyl ether			0.567	0.587	0.534	0.576	0.568	0.561		0.549	0.563	3.077		
T	Carbon Tetrachloride		0.444	0.456	0.462	0.441	0.481	0.465	0.474	0.462	0.448	0.4594	2.92173		
S	1,2-Dichloroethane-d4			0.241	0.309	0.276	0.312	0.295	0.301	0.284	0.286	0.288	7.93644		
T	Heptane											0	0		
T	1,2-Dichloroethane		0.347	0.335	0.354	0.336	0.371	0.36	0.364	0.349	0.338	0.3505	3.65503		
T	Benzene		1.128	1.084	1.071	0.977	1.085	1.033	1.019	0.963	0.883	1.0269	7.39343		
T	Trichloroethene		0.285	0.275	0.276	0.251	0.283	0.273	0.279	0.273	0.27	0.2739	3.66126		
T	Methylcyclohexane					0.396	0.426	0.408	0.429	0.416	0.422	0.4161	2.94153		
C	1,2-Dichloropropane		0.239	0.222	0.233	0.207	0.238	0.226	0.233	0.231	0.229	0.2286	4.30096		
T	Bromodichloromethane		0.369	0.382	0.412	0.385	0.426	0.421	0.417	0.41	0.395	0.4018	4.93305		
T	1,4-Dioxane					0.001	0.001	0.001	0.001		0.001	0.0011	14.5817		
T	Dibromomethane		0.141	0.13	0.137	0.128	0.141	0.134	0.137	0.136	0.133	0.1353	3.26799		
T	2-Chloroethyl Vinyl Ether				0.1	0.084	0.103	0.104	0.103	0.095	0.098	0.0983	7.09792		
T	4-Methyl-2-Pentanone					0.033	0.042	0.042	0.043	0.04	0.041	0.0401	8.66859		
T	cis-1,3-Dichloropropene		0.391	0.393	0.386	0.398	0.431	0.428	0.428	0.42	0.405	0.409	4.41527		

T	Dimethyl Disulfide					0.182	0.218	0.221	0.23	0.229	0.23	0.2183	8.49419	
I	Chlorobenzene-d5	ISTD												
S	Toluene-d8			1.282	1.295	1.228	1.385	1.333	1.378	1.289	1.278	1.3084	4.07746	
C	Toluene		1.653	1.696	1.678	1.507	1.647	1.554	1.545	1.418	1.27	1.552	8.98325	
T	Ethyl Methacrylate			0.251	0.293	0.268	0.314	0.314	0.327	0.309	0.31	0.2981	8.76897	
T	Paraldehyde											0	0	
T	trans-1,3-Dichloropropene			0.472	0.51	0.466	0.561	0.552	0.553	0.533	0.515	0.5201	7.01006	
T	1,1,2-Trichloroethane		0.261	0.241	0.242	0.222	0.254	0.249	0.252	0.243	0.241	0.2449	4.54209	
T	2-Hexanone					0.042	0.053	0.053	0.056	0.051	0.053	0.0515	9.50479	
T	1,3-Dichloropropane		0.355	0.442	0.465	0.42	0.468	0.454	0.457	0.437	0.428	0.436	7.93348	
T	Tetrachloroethene		0.355	0.337	0.336	0.306	0.333	0.325	0.337	0.334	0.33	0.3325	3.83886	
T	Dibromochloromethane		0.349	0.336	0.338	0.328	0.373	0.372	0.383	0.373	0.367	0.3576	5.60043	
T	1,2-Dibromoethane		0.233	0.234	0.234	0.227	0.259	0.254	0.258	0.251	0.247	0.2441	5.04151	
T	1-Chlorohexane		0.532	0.584	0.505	0.502	0.54	0.52	0.547	0.521	0.516	0.5298	4.7826	
P	Chlorobenzene		1.029	1.173	1.09	1.007	1.099	1.037	1.043	1.002	0.939	1.0466	6.41465	
T	1,1,1,2-Tetrachloroethane		0.392	0.399	0.403	0.357	0.431	0.414	0.424	0.417	0.413	0.4055	5.39495	
C	Ethylbenzene		0.6	0.594	0.59	0.532	0.586	0.559	0.57	0.564	0.558	0.5726	3.7769	
T	m-,p-Xylene		0.744	0.696	0.691	0.643	0.713	0.67	0.673	0.634	0.581	0.6716	7.12586	
T	o-Xylene			0.691	0.651	0.614	0.692	0.658	0.667	0.65	0.63	0.6567	4.11758	
T	Styrene		1.05	1.06	1.096	1.043	1.159	1.124	1.12	1.061	0.98	1.077	4.96781	
P	Bromoform			0.177	0.21	0.181	0.226	0.221	0.232	0.227	0.223	0.2121	10.0546	
T	Isopropylbenzene		1.853	1.891	1.871	1.708	1.897	1.781	1.75	1.573	1.377	1.7444	9.92447	
I	1,4-Dichlorobenzene-d4	ISTD												
P	1,1,2,2-Tetrachloroethane		0.426	0.492	0.503	0.486	0.543	0.51	0.527	0.513	0.514	0.5018	6.60291	
S	p-Bromofluorobenzene				1.054	0.976	1.071	1.016	1.053	1.031	1.046	1.0352	3.05136	
T	1,2,3-Trichloropropane			0.119	0.158	0.161	0.162	0.155	0.162	0.16	0.158	0.1545	9.40789	
T	trans-1,4-Dichloro-2-Butene			0.085	0.118	0.12	0.149	0.151	0.16	0.159	0.161	0.1376	19.9543	1.000
T	n-Propylbenzene		4.139	4.239	4.243	4.051	4.241	3.953	3.882	3.419	2.935	3.9003	11.4439	
T	Bromobenzene	0.796	0.708	0.843	0.837	0.778	0.84	0.798	0.817	0.803	0.803	0.8022	4.93253	
T	1,3,5-Trimethylbenzene		2.941	2.93	2.941	2.839	3.053	2.882	2.871	2.689	2.423	2.8411	6.50391	
T	2-Chlorotoluene		2.917	2.935	2.929	2.712	2.861	2.795	2.682	2.533	2.145	2.7231	9.37022	
T	4-Chlorotoluene		2.681	2.596	2.581	2.474	2.625	2.326	2.447	2.286	2.121	2.4597	7.49558	
T	a-Methylstyrene					1.257	1.406	1.34	1.384	1.35	1.351	1.3479	3.77654	
T	tert-Butylbenzene			0.587	0.555	0.53	0.571	0.533	0.552	0.561	0.558	0.5558	3.34508	
T	1,2,4-Trimethylbenzene		2.955	3.168	3.127	2.936	3.12	2.914	2.937	2.709	2.427	2.9214	7.97314	
T	sec-Butylbenzene			3.701	3.536	3.487	3.607	3.34	3.332	3.045	2.69	3.3424	9.93324	
T	p-Isopropyltoluene			3.112	3.062	2.903	2.988	2.813	2.809	2.634	2.382	2.8379	8.45265	
T	1,3-Dichlorobenzene		1.484	1.617	1.627	1.53	1.605	1.531	1.549	1.51	1.45	1.5448	3.95774	
T	1,4-Dichlorobenzene	1.713	1.524	1.672	1.594	1.503	1.6	1.511	1.535	1.49	1.431	1.5574	5.56562	
T	n-Butylbenzene			3.143	2.96	2.895	3.002	2.791	2.792	2.571	2.292	2.8056	9.55812	
T	1,2-Dichlorobenzene	1.59	1.307	1.415	1.41	1.322	1.404	1.333	1.356	1.315	1.286	1.3739	6.4618	
T	1,2-Dibromo-3-Chloropropane				0.085	0.103	0.102	0.103	0.106	0.104	0.103	0.1007	6.86771	
T	1,2,4-Trichlorobenzene		1.074	1.01	1.028	0.982	1.069	1.013	1.036	1.029	1.002	1.0269	2.90829	
T	Hexachlorobutadiene		0.516	0.476	0.483	0.447	0.487	0.459	0.48	0.485	0.485	0.4798	4.00465	
T	Naphthalene		1.594	1.457	1.574	1.455	1.63	1.552	1.564	1.485	1.422	1.5258	4.75891	
T	1,2,3-Trichlorobenzene	0.913	0.934	0.905	0.862	0.805	0.877	0.824	0.862	0.846	0.835	0.8664	4.7474	

Tue May 17 11:49:41 2016

Calibration Table Report
Method: A9FOOWT.M
Title: A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Calibration: Thu May 19 08:54:31 2016
Curve: WG569079
Calibration Files

	Compound	5 8M412324.D	20 8M412323.D	50 8M412322.D	100 8M412321.D	200 8M412325.D	300 8M412326.D	400 8M412327.D	500 8M412328.D	Avg	%RSD
I	Fluorobenzene	ISTD									
T	Acetonitrile		0.009	0.010	0.010	0.010	0.010	0.010	0.010	0.010	2.076
T	3-Chloro-1-propene	0.306	0.330	0.335	0.331	0.324	0.307	0.304	0.304	0.318	4.307
T	2-Chloro-1,3-butadiene	0.368	0.420	0.420	0.419	0.419	0.392	0.389	0.387	0.402	5.022
T	Ethyl Acetate	0.080	0.096	0.097	0.100	0.099	0.098	0.100	0.095	0.096	6.751
T	Methacrylonitrile	0.041	0.054	0.055	0.055	0.055	0.054	0.055	0.052	0.053	9.227
T	Isobutyl Alcohol			0.003	0.003	0.003	0.003	0.003	0.003	0.003	3.599
T	1-Butanol									0.000	0.000
T	Methyl methacrylate	0.089	0.106	0.106	0.105	0.106	0.105	0.107	0.102	0.103	5.868
T	2-Nitropropane		0.041	0.042	0.043	0.046	0.047	0.048	0.046	0.045	5.992
I	Chlorobenzene-d5	ISTD									
I	1,4-Dichlorobenzene-d4	ISTD									
T	Cyclohexanone		0.011	0.010	0.010	0.009	0.010	0.011	0.010	0.010	4.971

Thu May 19 09:04:52 2016

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16060823 Run Date: 05/13/2016 Sample ID: WG568561-12
Instrument ID: HPMS8 Run Time: 19:17 Method: 8260B
File ID: 8M412232 Analyst: TMB QC Key: DOD4
ICal Workgroup: WG568561 Cal ID: HPMS8 - 13-MAY-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
Chloroform	CCC	50.0	47.1	ug/L	0.513	5.70	20	
1,1-Dichloroethene	CCC	50.0	46.0	ug/L	0.372	8.10	20	
1,2-Dichloropropane	CCC	50.0	51.0	ug/L	0.233	1.90	20	
Ethylbenzene	CCC	50.0	48.2	ug/L	0.552	3.60	20	
Toluene	CCC	50.0	49.1	ug/L	1.52	1.80	20	
Vinyl Chloride	CCC	50.0	52.4	ug/L	0.304	4.80	20	
Bromoform	SPCC	50.0	48.2	ug/L	0.205	3.60	20	
Chlorobenzene	SPCC	50.0	50.2	ug/L	1.05	0.400	20	
Chloromethane	SPCC	50.0	50.7	ug/L	0.248	1.40	20	
1,1-Dichloroethane	SPCC	50.0	46.7	ug/L	0.431	6.60	20	
1,1,2,2-Tetrachloroethane	SPCC	50.0	49.6	ug/L	0.498	0.800	20	
Acetone		50.0	57.5	ug/L	0.0324	15.1	20	
Benzene		50.0	48.5	ug/L	0.997	3.00	20	
Bromobenzene		50.0	49.5	ug/L	0.794	1.10	20	
Bromochloromethane		50.0	47.5	ug/L	0.140	4.90	20	
Bromodichloromethane		50.0	48.3	ug/L	0.388	3.40	20	
Bromomethane		50.0	46.4	ug/L	0.162	7.20	20	
2-Butanone		50.0	53.0	ug/L	0.0438	6.00	20	
n-Butylbenzene		50.0	49.7	ug/L	2.79	0.700	20	
sec-Butylbenzene		50.0	51.3	ug/L	3.43	2.70	20	
tert-Butylbenzene		50.0	50.9	ug/L	0.566	1.90	20	
Carbon Disulfide		50.0	42.3	ug/L	0.716	15.4	20	
Carbon Tetrachloride		50.0	46.2	ug/L	0.425	7.50	20	
Dibromochloromethane		50.0	49.7	ug/L	0.355	0.700	20	
Chloroethane		50.0	53.3	ug/L	0.170	6.70	20	
2-Chlorotoluene		50.0	51.6	ug/L	2.81	3.10	20	
4-Chlorotoluene		50.0	48.3	ug/L	2.37	3.50	20	
1,2-Dibromo-3-Chloropropane		50.0	48.5	ug/L	0.0977	3.00	20	
1,2-Dibromoethane		50.0	51.1	ug/L	0.250	2.20	20	
Dibromomethane		50.0	45.8	ug/L	0.124	8.40	20	
1,2-Dichlorobenzene		50.0	50.5	ug/L	1.39	1.00	20	
1,3-Dichlorobenzene		50.0	50.9	ug/L	1.57	1.80	20	
1,4-Dichlorobenzene		50.0	50.5	ug/L	1.57	0.900	20	
Dichlorodifluoromethane		50.0	54.2	ug/L	0.455	8.40	20	
1,2-Dichloroethane		50.0	49.0	ug/L	0.344	2.00	20	
cis-1,2-Dichloroethene		50.0	48.9	ug/L	0.281	2.20	20	
trans-1,2-Dichloroethene		50.0	48.1	ug/L	0.357	3.90	20	
1,3-Dichloropropane		50.0	53.7	ug/L	0.468	7.30	20	
2,2-Dichloropropane		50.0	47.9	ug/L	0.466	4.20	20	
cis-1,3-Dichloropropene		50.0	53.9	ug/L	0.441	7.80	20	
trans-1,3-Dichloropropene		50.0	50.7	ug/L	0.527	1.30	20	
1,1-Dichloropropene		50.0	47.5	ug/L	0.374	5.10	20	

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4819686
Report generated 07/07/2016 16:44

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16060823 Run Date: 05/13/2016 Sample ID: WG568561-12
Instrument ID: HPMS8 Run Time: 19:17 Method: 8260B
File ID: 8M412232 Analyst: TMB QC Key: DOD4
ICal Workgroup: WG568561 Cal ID: HPMS8 - 13-MAY-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
2-Hexanone		50.0	53.6	ug/L	0.0552	7.20	20	
Hexachlorobutadiene		50.0	52.1	ug/L	0.500	4.20	20	
Isopropylbenzene		50.0	51.1	ug/L	1.78	2.20	20	
p-Isopropyltoluene		50.0	50.1	ug/L	2.85	0.300	20	
4-Methyl-2-Pentanone		50.0	51.9	ug/L	0.0416	3.70	20	
Methylene Chloride		50.0	48.4	ug/L	0.249	3.10	20	
Naphthalene		50.0	45.8	ug/L	1.40	8.40	20	
n-Propylbenzene		50.0	51.9	ug/L	4.05	3.80	20	
Styrene		50.0	51.5	ug/L	1.11	3.00	20	
1,1,1,2-Tetrachloroethane		50.0	49.4	ug/L	0.400	1.20	20	
Tetrachloroethene		50.0	47.8	ug/L	0.318	4.40	20	
1,2,3-Trichlorobenzene		50.0	47.7	ug/L	0.827	4.50	20	
1,2,4-Trichlorobenzene		50.0	50.1	ug/L	1.03	0.200	20	
1,1,1-Trichloroethane		50.0	47.6	ug/L	0.492	4.80	20	
1,1,2-Trichloroethane		50.0	49.0	ug/L	0.240	1.90	20	
Trichloroethene		50.0	51.1	ug/L	0.280	2.10	20	
Trichlorofluoromethane		50.0	46.9	ug/L	0.507	6.10	20	
1,2,3-Trichloropropane		50.0	50.1	ug/L	0.155	0.200	20	
1,2,4-Trimethylbenzene		50.0	49.7	ug/L	2.90	0.700	20	
1,3,5-Trimethylbenzene		50.0	51.6	ug/L	2.93	3.30	20	
o-Xylene		50.0	50.6	ug/L	0.664	1.10	20	
m-,p-Xylene		100	101	ug/L	0.679	1.10	20	

* Exceeds %D Limit

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4819686
Report generated 07/07/2016 16:44

Microbac Laboratories Inc.
CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823 Run Date: 06/16/2016 Sample ID: WG573037-02
Instrument ID: HPMS8 Run Time: 14:25 Method: 8260B
File ID: 8M413032 Analyst: FJB QC Key: DOD4
Workgroup (AAB#): WG573038 Cal ID: HPMS8 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Chloroform	CCC	50.0	54.8	ug/L	0.597	9.65	20	
1,1-Dichloroethene	CCC	50.0	54.7	ug/L	0.442	9.30	20	
1,2-Dichloropropane	CCC	50.0	48.6	ug/L	0.222	2.70	20	
Ethylbenzene	CCC	50.0	50.5	ug/L	0.578	0.905	20	
Toluene	CCC	50.0	50.3	ug/L	1.56	0.600	20	
Vinyl Chloride	CCC	50.0	55.1	ug/L	0.320	10.2	20	
Bromoform	SPCC	50.0	50.9	ug/L	0.216	1.72	20	
Chlorobenzene	SPCC	50.0	49.8	ug/L	1.04	0.474	20	
Chloromethane	SPCC	50.0	46.6	ug/L	0.228	6.88	20	
1,1-Dichloroethane	SPCC	50.0	52.3	ug/L	0.483	4.61	20	
1,1,2,2-Tetrachloroethane	SPCC	50.0	41.5	ug/L	0.416	17.0	20	
Xylenes		150	154	ug/L	0.683	2.90	20	
Acetone		50.0	52.2	ug/L	0.0293	4.31	20	
Benzene		50.0	52.0	ug/L	1.07	3.99	20	
Bromobenzene		50.0	48.0	ug/L	0.770	4.08	20	
Bromochloromethane		50.0	51.3	ug/L	0.151	2.54	20	
Bromodichloromethane		50.0	55.4	ug/L	0.445	10.8	20	
Bromomethane		50.0	37.0	ug/L	0.128	26.0	20	*
2-Butanone		50.0	45.9	ug/L	0.0380	8.13	20	
n-Butylbenzene		50.0	50.5	ug/L	2.83	1.00	20	
sec-Butylbenzene		50.0	50.3	ug/L	3.37	0.699	20	
tert-Butylbenzene		50.0	48.9	ug/L	0.544	2.12	20	
Carbon Disulfide		50.0	53.0	ug/L	0.897	6.05	20	
Carbon Tetrachloride		50.0	60.0	ug/L	0.551	20.0	20	
Dibromochloromethane		50.0	50.5	ug/L	0.361	0.983	20	
Chloroethane		50.0	48.9	ug/L	0.156	2.14	20	
2-Chlorotoluene		50.0	47.7	ug/L	2.60	4.50	20	
4-Chlorotoluene		50.0	49.7	ug/L	2.44	0.615	20	
1,2-Dibromo-3-Chloropropane		50.0	44.0	ug/L	0.0886	12.0	20	
1,2-Dibromoethane		50.0	47.2	ug/L	0.230	5.67	20	
Dibromomethane		50.0	49.7	ug/L	0.135	0.584	20	
1,2-Dichlorobenzene		50.0	46.6	ug/L	1.28	6.76	20	
1,3-Dichlorobenzene		50.0	49.6	ug/L	1.53	0.820	20	
1,4-Dichlorobenzene		50.0	47.6	ug/L	1.48	4.82	20	
Dichlorodifluoromethane		50.0	57.6	ug/L	0.483	15.3	20	
1,2-Dichloroethane		50.0	56.7	ug/L	0.397	13.3	20	
cis-1,2-Dichloroethene		50.0	50.9	ug/L	0.292	1.90	20	
trans-1,2-Dichloroethene		50.0	52.1	ug/L	0.387	4.10	20	
1,3-Dichloropropane		50.0	47.0	ug/L	0.410	5.96	20	
2,2-Dichloropropane		50.0	61.8	ug/L	0.601	23.6	20	*
cis-1,3-Dichloropropene		50.0	53.3	ug/L	0.436	6.58	20	
trans-1,3-Dichloropropene		50.0	51.1	ug/L	0.532	2.30	20	

CCV - Modified 03/05/2008
PDF File ID: 4819688
Report generated 07/07/2016 16:44

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823 Run Date: 06/16/2016 Sample ID: WG573037-02
Instrument ID: HPMS8 Run Time: 14:25 Method: 8260B
File ID: 8M413032 Analyst: FJB QC Key: DOD4
Workgroup (AAB#): WG573038 Cal ID: HPMS8 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
1,1-Dichloropropene		50.0	54.6	ug/L	0.430	9.13	20	
2-Hexanone		50.0	44.4	ug/L	0.0457	11.3	20	
Hexachlorobutadiene		50.0	54.7	ug/L	0.525	9.43	20	
Isopropylbenzene		50.0	54.0	ug/L	1.88	7.91	20	
p-Isopropyltoluene		50.0	50.6	ug/L	2.87	1.20	20	
4-Methyl-2-Pentanone		50.0	45.5	ug/L	0.0366	8.93	20	
Methylene Chloride		50.0	48.8	ug/L	0.251	2.36	20	
Naphthalene		50.0	43.5	ug/L	1.33	13.0	20	
n-Propylbenzene		50.0	49.8	ug/L	3.88	0.484	20	
Styrene		50.0	52.0	ug/L	1.12	4.02	20	
1,1,1,2-Tetrachloroethane		50.0	52.3	ug/L	0.424	4.56	20	
Tetrachloroethene		50.0	53.6	ug/L	0.356	7.11	20	
1,2,3-Trichlorobenzene		50.0	45.4	ug/L	0.786	9.28	20	
1,2,4-Trichlorobenzene		50.0	48.6	ug/L	0.997	2.89	20	
1,1,1-Trichloroethane		50.0	59.6	ug/L	0.616	19.1	20	
1,1,2-Trichloroethane		50.0	46.3	ug/L	0.227	7.39	20	
Trichloroethene		50.0	54.0	ug/L	0.296	7.92	20	
Trichlorofluoromethane		50.0	59.7	ug/L	0.645	19.5	20	
1,2,3-Trichloropropane		50.0	45.0	ug/L	0.139	10.0	20	
1,2,4-Trimethylbenzene		50.0	50.1	ug/L	2.93	0.254	20	
1,3,5-Trimethylbenzene		50.0	50.2	ug/L	2.85	0.347	20	
o-Xylene		50.0	51.2	ug/L	0.673	2.46	20	
m-,p-Xylene		100	103	ug/L	0.693	3.11	20	

* Exceeds %D Criteria

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

CCV - Modified 03/05/2008
PDF File ID: 4819688
Report generated 07/07/2016 16:44

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060823
Instrument ID: HPMS8
Workgroup (AAB#): WG573038

ICAL CCV Number: WG568561-08
CAL ID: HPMS8 - 13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568561-08	NA	NA	288225	539116	778260
Upper Limit	NA	NA	576450	1078232	1556520
Lower Limit	NA	NA	144113	269558	389130
L16060823-01	1.00	01	295882	526044	672547
L16060823-03	1.00	01	289710	517845	676219
L16060823-04	1.00	01	287793	517811	671491
L16060823-05	1.00	01	289040	525559	704722
WG573038-01	1.00	01	292043	533754	717429
WG573038-02	1.00	01	307888	540053	724153
WG573038-03	1.00	01	312159	553122	735322

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_ICAL - Modified 03/06/2008
PDF File ID: 4819689
Report generated 07/07/2016 16:44

Microbac Laboratories Inc.

INTERNAL STANDARD RETENTION TIME SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060823
Instrument ID: HPMS8
Workgroup (AAB#): WG573038

ICAL CCV Number: WG568561-08
CAL ID: HPMS8-13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568561-08	NA	NA	17.6	14.58	10.71
Upper Limit	NA	NA	18.1	15.08	11.21
Lower Limit	NA	NA	17.1	14.08	10.21
L16060823-01	1.00	01	17.6	14.57	10.71
L16060823-03	1.00	01	17.6	14.58	10.71
L16060823-04	1.00	01	17.6	14.57	10.71
L16060823-05	1.00	01	17.6	14.58	10.71
WG573038-01	1.00	01	17.6	14.57	10.71
WG573038-02	1.00	01	17.6	14.57	10.71
WG573038-03	1.00	01	17.6	14.57	10.71

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_RT_ICAL - Modified 03/06/2008
PDF File ID: 4819691
Report generated: 07/07/2016 16:44

2.1.1.3 Sample Data

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413045.D Vial: 15
Acq On : 16 Jun 2016 20:50 Operator: FJB
Sample : L16060823-01 A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 08:18:24 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	672547	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	526044	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	295882	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	205604	28.4536 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 113.80%		
43) 1,2-Dichloroethane-d4	10.30	65	234064	30.2095 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 120.84%#		
58) Toluene-d8	12.68	98	698178	25.3600 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 101.44%		
80) p-Bromofluorobenzene	16.08	95	304815	24.8796 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 99.52%		

Target Compounds					Qvalue
3) Chloromethane	3.60	50	937	0.1424 ug/L	80
47) Trichloroethene	11.21	130	1071	0.1454 ug/L #	72
92) p-Isopropyltoluene	17.32	119	42966	1.2792 ug/L	97

(#) = qualifier out of range (m) = manual integration
8M413045.D 8260WT.M Mon Jun 20 09:20:05 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413045.D Vial: 15
Acq On : 16 Jun 2016 20:50 Operator: FJB
Sample : L16060823-01 A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 8:18 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413045.D 8260WT.M Mon Jun 20 09:20:06 2016 Page 2



#3
Chloromethane
Concen: 0.14 ug/L
RT: 3.60 min Scan# 127
Delta R.T. -0.02 min
Lab File: 8M413045.D
Acq: 16 Jun 2016 20:50

Tgt Ion: 50 Resp: 937

Ion	Ratio	Lower	Upper
50	100		
52	21.7	19.9	46.3





#47
Trichloroethene
Concen: 0.15 ug/L
RT: 11.21 min Scan# 864
Delta R.T. -0.00 min
Lab File: 8M413045.D
Acq: 16 Jun 2016 20:50

Tgt Ion:130 Resp: 1071

Ion	Ratio	Lower	Upper
130	100		
132	48.3	58.0	135.2#
95	117.9	64.9	151.5





#92
p-Isopropyltoluene
Concen: 1.28 ug/L
RT: 17.32 min Scan# 1455
Delta R.T. -0.00 min
Lab File: 8M413045.D
Acq: 16 Jun 2016 20:50

Tgt Ion:119 Resp: 42966

Ion	Ratio	Lower	Upper
119	100		
134	23.9	15.2	35.6



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413046.D Vial: 16
Acq On : 16 Jun 2016 21:19 Operator: FJB
Sample : L16060823-03 A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 08:18:25 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	676219	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	517845	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	289710	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	203536	28.0144 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 112.04%		
43) 1,2-Dichloroethane-d4	10.30	65	232456	29.8391 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 119.36%		
58) Toluene-d8	12.68	98	703041	25.9409 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 103.76%		
80) p-Bromofluorobenzene	16.08	95	300473	25.0476 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 100.20%		
Target Compounds					Qvalue
3) Chloromethane	3.62	50	1964	0.2968 ug/L #	42
32) cis-1,2-Dichloroethene	9.17	96	1645	0.2119 ug/L	72
47) Trichloroethene	11.21	130	41569	5.6110 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M413046.D 8260WT.M Mon Jun 20 09:20:10 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413046.D Vial: 16
Acq On : 16 Jun 2016 21:19 Operator: FJB
Sample : L16060823-03 A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 8:18 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413046.D 8260WT.M Mon Jun 20 09:20:10 2016 Page 2



#3
Chloromethane
Concen: 0.30 ug/L
RT: 3.62 min Scan# 129
Delta R.T. -0.00 min
Lab File: 8M413046.D
Acq: 16 Jun 2016 21:19

Tgt Ion: 50 Resp: 1964

Ion	Ratio	Lower	Upper
50	100		
52	0.0	19.9	46.3#





#32
cis-1,2-Dichloroethene
Concen: 0.21 ug/L
RT: 9.17 min Scan# 666
Delta R.T. -0.00 min
Lab File: 8M413046.D
Acq: 16 Jun 2016 21:19

Tgt Ion: 96 Resp: 1645

Ion	Ratio	Lower	Upper
96	100		
61	116.5	91.5	213.5





#47
Trichloroethene
Concen: 5.61 ug/L
RT: 11.21 min Scan# 864
Delta R.T. -0.00 min
Lab File: 8M413046.D
Acq: 16 Jun 2016 21:19

Tgt Ion:130 Resp: 41569

Ion	Ratio	Lower	Upper
130	100		
132	96.8	58.0	135.2
95	106.6	64.9	151.5



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413047.D Vial: 17
Acq On : 16 Jun 2016 21:48 Operator: FJB
Sample : L16060823-04 A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 08:18:25 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	671491	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	517811	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	287793	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	203988	28.2743 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 113.08%		
43) 1,2-Dichloroethane-d4	10.30	65	233925	30.2391 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 120.96%#		
58) Toluene-d8	12.68	98	705263	26.0246 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 104.08%		
80) p-Bromofluorobenzene	16.08	95	298621	25.0591 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 100.24%		

Target Compounds					Qvalue
3) Chloromethane	3.63	50	1302	0.1981 ug/L #	42

(#) = qualifier out of range (m) = manual integration
8M413047.D 8260WT.M Mon Jun 20 09:20:14 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413047.D Vial: 17
Acq On : 16 Jun 2016 21:48 Operator: FJB
Sample : L16060823-04 A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 8:18 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413047.D 8260WT.M Mon Jun 20 09:20:14 2016 Page 2



#3
Chloromethane
Concen: 0.20 ug/L
RT: 3.63 min Scan# 130
Delta R.T. 0.01 min
Lab File: 8M413047.D
Acq: 16 Jun 2016 21:48

Tgt Ion: 50 Resp: 1302

Ion	Ratio	Lower	Upper
50	100		
52	0.0	19.9	46.3#



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413042.D Vial: 12
Acq On : 16 Jun 2016 19:23 Operator: FJB
Sample : L16060823-05 TB A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 08:18:24 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	704722	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	525559	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	289040	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	208489	27.5355 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 110.16%		
43) 1,2-Dichloroethane-d4	10.30	65	231607	28.5277 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 114.12%		
58) Toluene-d8	12.68	98	729703	26.5295 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 106.12%		
80) p-Bromofluorobenzene	16.08	95	306574	25.6155 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 102.48%		

Target Compounds					Qvalue
36) Tetrahydrofuran	9.65	42	1074	1.3998 ug/L #	56

(#) = qualifier out of range (m) = manual integration
8M413042.D 8260WT.M Mon Jun 20 09:19:51 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413042.D Vial: 12
Acq On : 16 Jun 2016 19:23 Operator: FJB
Sample : L16060823-05 TB A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 8:18 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413042.D 8260WT.M Mon Jun 20 09:19:52 2016 Page 2



#36
Tetrahydrofuran
Concen: 1.40 ug/L
RT: 9.65 min Scan# 713
Delta R.T. 0.02 min
Lab File: 8M413042.D
Acq: 16 Jun 2016 19:23

Tgt Ion: 42 Resp: 1074

Ion	Ratio	Lower	Upper
42	100		
72	27.6	33.8	78.8#
71	18.9	32.0	74.8#



2.1.1.4 Standards Data

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\2\DATA\051316\8M412220.D Vial: 3
Acq On : 13 May 2016 13:29 Operator: TMB
Sample : WG568561-02 0.3ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 17 11:40:52 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	823748	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	571926	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	299321	25.00 ug/L	0.00

System Monitoring Compounds					
37) Dibromofluoromethane	0.00	111	0	0.0000 ug/L	
Spiked Amount 25.000	Range 86 - 118		Recovery	= 0.00%#	
43) 1,2-Dichloroethane-d4	0.00	65	0	0.0000 ug/L	
Spiked Amount 25.000	Range 80 - 120		Recovery	= 0.00%#	
58) Toluene-d8	12.68	98	567	0.0189 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery	= 0.08%#	
80) p-Bromofluorobenzene	16.08	95	313	0.0253 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery	= 0.12%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.17	85	3836	0.2778 ug/L	91
3) Chloromethane	3.62	50	3658	0.4537 ug/L	95
4) Vinyl Chloride	3.85	62	2516	0.2630 ug/L #	58
5) 1,3-Butadiene	3.91	54	2079	0.2592 ug/L	71
6) Bromomethane	4.73	94	1152	0.6530 ug/L #	55
7) Chloroethane	4.90	64	1112	0.2113 ug/L #	72
8) Trichlorofluoromethane	5.35	101	2706	0.1520 ug/L #	53
10) Isoprene	5.93	67	2147	0.1987 ug/L	85
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	1746	0.2087 ug/L	97
14) 1,1-Dichloroethene	6.48	61	3332	0.2501 ug/L	75
16) Dimethyl Sulfide	6.73	62	1099	0.2308 ug/L	95
19) Methylene Chloride	7.26	84	3414	0.4028 ug/L	88
20) Carbon Disulfide	7.30	76	10739	0.3853 ug/L	97
22) Methyl Tert Butyl Ether	7.47	73	4927	0.2686 ug/L #	60
23) trans-1,2-Dichloroethene	7.70	61	3879	0.3170 ug/L	98
24) n-Hexane	7.78	57	2794	0.3075 ug/L	85
26) Vinyl Acetate	8.30	43	1973	0.2259 ug/L #	70
27) 1,1-Dichloroethane	8.33	63	4881	0.3212 ug/L #	88
31) 2,2-Dichloropropane	9.11	77	4966	0.3101 ug/L	89
32) cis-1,2-Dichloroethene	9.17	96	2958	0.3128 ug/L	87
33) Chloroform	9.37	83	5564	0.3101 ug/L	95
34) 1-Bromopropane	9.52	122	210	0.1551 ug/L #	1
35) Bromochloromethane	9.60	130	1386	0.2858 ug/L	76
36) Tetrahydrofuran	9.62	42	1849	2.0617 ug/L #	60
38) 1,1,1-Trichloroethane	9.91	97	4978	0.2921 ug/L	95
39) Cyclohexane	9.93	56	3120	0.2644 ug/L #	85
40) 1,1-Dichloropropene	10.11	75	3779	0.2909 ug/L	97
42) Carbon Tetrachloride	10.24	117	4835	0.3194 ug/L #	74
45) 1,2-Dichloroethane	10.43	62	3295	0.2853 ug/L #	81
46) Benzene	10.46	78	11279	0.3333 ug/L	93
47) Trichloroethene	11.20	130	2410	0.2670 ug/L	97
48) Methylcyclohexane	11.31	83	3936	0.2871 ug/L	84
49) 1,2-Dichloropropane	11.42	63	2136	0.2836 ug/L	88
50) Bromodichloromethane	11.73	83	3767	0.2845 ug/L #	91
52) Dibromomethane	11.82	93	700	0.1570 ug/L #	59
53) 2-Chloroethyl Vinyl Ether	12.02	63	847	0.2616 ug/L #	47
55) cis-1,3-Dichloropropene	12.36	75	4763	0.3535 ug/L	84
56) Dimethyl Disulfide	12.63	79	1647	0.2290 ug/L	82
59) Toluene	12.78	91	11513	0.3243 ug/L	92
60) Ethyl Methacrylate	12.89	69	1631	0.2392 ug/L #	64

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\2\DATA\051316\8M412220.D Vial: 3
Acq On : 13 May 2016 13:29 Operator: TMB
Sample : WG568561-02 0.3ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 17 11:40:52 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
62)	trans-1,3-Dichloropropene	12.96	75	3626	0.3047 ug/L		82
63)	1,1,2-Trichloroethane	13.18	97	1409	0.2515 ug/L		98
65)	1,3-Dichloropropane	13.49	76	2584	0.2590 ug/L		93
66)	Tetrachloroethene	13.61	164	2678	0.3520 ug/L	#	75
67)	Dibromochloromethane	13.87	129	2126	0.2599 ug/L		79
68)	1,2-Dibromoethane	14.12	107	1895	0.3394 ug/L		91
69)	1-Chlorohexane	14.21	91	3835	0.3164 ug/L		99
70)	Chlorobenzene	14.62	112	7968	0.3328 ug/L		88
71)	1,1,1,2-Tetrachloroethane	14.67	131	2872	0.3096 ug/L		87
72)	Ethylbenzene	14.64	106	3844	0.2934 ug/L		81
73)	m-,p-Xylene	14.75	106	10447	0.6799 ug/L		96
74)	o-Xylene	15.31	106	4577	0.3047 ug/L		97
75)	Styrene	15.35	104	7545	0.3062 ug/L		95
76)	Bromoform	15.84	173	1301	0.2681 ug/L	#	51
77)	Isopropylbenzene	15.74	105	13119	0.3287 ug/L		90
79)	1,1,2,2-Tetrachloroethane	15.96	83	1714	0.2853 ug/L	#	85
83)	n-Propylbenzene	16.24	91	15676	0.3357 ug/L		92
84)	Bromobenzene	16.38	156	2859	0.2977 ug/L		90
85)	1,3,5-Trimethylbenzene	16.43	105	10873	0.3196 ug/L		99
86)	2-Chlorotoluene	16.52	91	11133	0.3415 ug/L		96
87)	4-Chlorotoluene	16.56	91	10305	0.3499 ug/L		97
88)	a-Methylstyrene	16.83	118	4090	0.2534 ug/L		86
89)	tert-Butylbenzene	16.90	134	1734	0.2606 ug/L		72
90)	1,2,4-Trimethylbenzene	16.95	105	10734	0.3069 ug/L		89
91)	sec-Butylbenzene	17.16	105	13749	0.3436 ug/L		97
92)	p-Isopropyltoluene	17.32	119	10847	0.3192 ug/L		95
93)	1,3-Dichlorobenzene	17.52	146	6023	0.3257 ug/L		99
94)	1,4-Dichlorobenzene	17.65	146	6152	0.3299 ug/L	#	60
95)	n-Butylbenzene	17.85	91	11161	0.3323 ug/L		97
96)	1,2-Dichlorobenzene	18.14	146	5711	0.3472 ug/L		92
98)	1,2,4-Trichlorobenzene	20.29	180	4203	0.3418 ug/L		91
99)	Hexachlorobutadiene	20.43	225	1894	0.3297 ug/L		84
100)	Naphthalene	20.65	128	6039	0.3306 ug/L		91
101)	1,2,3-Trichlorobenzene	20.96	180	3279	0.3161 ug/L		86

(#) = qualifier out of range (m) = manual integration
8M412220.D 8260WT.M Tue May 17 11:40:55 2016 Page 2

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412220.D Vial: 3
Acq On : 13 May 2016 13:29 Operator: TMB
Sample : WG568561-02 0.3ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:03:01 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	-1.0000	0.2778	0.0	100	0.00
3	P	Chloromethane	-1.0000	0.4537	0.0	100	0.00
4	C	Vinyl Chloride	-1.0000	0.2630	0.0	100	0.00
5	T	1,3-Butadiene	-1.0000	0.2592	0.0	100	0.03
6	T	Bromomethane	-1.0000	0.9911	0.0	100	0.00
7	T	Chloroethane	-1.0000	0.2113	0.0	100	0.00
8	T	Trichlorofluoromethane	-1.0000	0.1520	0.0	100	-0.02
9	T	Diethyl ether	-1.0000	0.0000	0.0	0	-5.90#
10	T	Isoprene	-1.0000	0.1987	0.0	100	0.00
11	T	Acrolein	-1.0000	0.0000	0.0	0	-6.13#
12	T	1,1,2-Trichloro-1,2,2-Trifl	-1.0000	0.2087	0.0	100	0.00
13	T	Acetone	-1.0000	0.0000	0.0	0	-6.25#
14	C	1,1-Dichloroethene	-1.0000	0.2501	0.0	100	0.02
15	T	Tert-Butyl Alcohol	-1.0000	0.0000	0.0	0	-6.58#
16	T	Dimethyl Sulfide	-1.0000	0.2308	0.0	100	0.00
17	T	Iodomethane	-1.0000	0.0000	0.0	0	-6.98#
18	T	Methyl acetate	-1.0000	0.0000	0.0	0	-7.00#
19	T	Methylene Chloride	-1.0000	0.4028	0.0	100	0.00
20	T	Carbon Disulfide	-1.0000	0.3853	0.0	100	0.00
21	T	Acrylonitrile	-1.0000	0.0000	0.0	0	-7.43#
22	T	Methyl Tert Butyl Ether	-1.0000	0.2686	0.0	100	0.00
23	T	trans-1,2-Dichloroethene	-1.0000	0.3170	0.0	100	0.00
24	T	n-Hexane	-1.0000	0.3075	0.0	100	0.00
25	T	Diisopropyl ether	-1.0000	0.0000	0.0	0	-8.13#
26	T	Vinyl Acetate	-1.0000	0.2259	0.0	100	0.00
27	P	1,1-Dichloroethane	-1.0000	0.3212	0.0	100	0.00
28	T	Ethyl-Tert-Butyl ether	-1.0000	0.0000	0.0	0	-8.71#
29	T	2-Butanone	-1.0000	0.0000	0.0	0	-8.88#
30	T	Propionitrile	-1.0000	0.0000	0.0	0	-8.98#
31	T	2,2-Dichloropropane	-1.0000	0.3101	0.0	100	0.00
32	T	cis-1,2-Dichloroethene	-1.0000	0.3128	0.0	100	0.00
33	C	Chloroform	0.3000	0.3101	-3.4	100	0.00
34		1-Bromopropane	-1.0000	0.1551	0.0	100	0.00
35	T	Bromochloromethane	-1.0000	0.2858	0.0	100	0.00
36	T	Tetrahydrofuran	-1.0000	2.0617	0.0	100	-0.01
37	S	Dibromofluoromethane	-1.0000	0.0000	0.0	0	-9.67#
38	T	1,1,1-Trichloroethane	-1.0000	0.2921	0.0	100	0.00
39	T	Cyclohexane	-1.0000	0.2644	0.0	100	0.00
40	T	1,1-Dichloropropene	-1.0000	0.2909	0.0	100	0.00
41	T	Tert-Amyl-Methyl ether	-1.0000	0.0000	0.0	0	-10.21#
42	T	Carbon Tetrachloride	-1.0000	0.3194	0.0	100	0.00
43	S	1,2-Dichloroethane-d4	-1.0000	0.0000	0.0	0	-10.31#
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	-1.0000	0.2853	0.0	100	0.00
46	T	Benzene	-1.0000	0.3333	0.0	100	0.00
47	T	Trichloroethene	-1.0000	0.2670	0.0	100	-0.01
48	T	Methylcyclohexane	-1.0000	0.2871	0.0	100	0.00
49	C	1,2-Dichloropropane	-1.0000	0.2836	0.0	100	0.00
50	T	Bromodichloromethane	-1.0000	0.2845	0.0	100	0.00
51	T	1,4-Dioxane	-1.0000	0.0000	0.0	0	-11.71#
52	T	Dibromomethane	-1.0000	0.1570	0.0	100	0.00
53	T	2-Chloroethyl Vinyl Ether	-1.0000	0.2616	0.0	100	0.00
54	T	4-Methyl-2-Pentanone	-1.0000	0.0000	0.0	0	-12.05#

(#) = Out of Range

8M412220.D 8260WT.M Sat May 14 18:03:33 2016 Page 1

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412220.D Vial: 3
Acq On : 13 May 2016 13:29 Operator: TMB
Sample : WG568561-02 0.3ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:03:01 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	-1.0000	0.3535	0.0	100	0.00
56	T	Dimethyl Disulfide	-1.0000	0.2290	0.0	100	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	-1.0000	0.0189	0.0	100	0.00
59	C	Toluene	-1.0000	0.3243	0.0	100	0.00
60	T	Ethyl Methacrylate	-1.0000	0.2392	0.0	100	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	-1.0000	0.3047	0.0	100	0.00
63	T	1,1,2-Trichloroethane	-1.0000	0.2515	0.0	100	0.00
64	T	2-Hexanone	-1.0000	0.0000	0.0	0	-13.12#
65	T	1,3-Dichloropropane	-1.0000	0.2590	0.0	100	0.00
66	T	Tetrachloroethene	-1.0000	0.3520	0.0	100	0.00
67	T	Dibromochloromethane	-1.0000	0.2599	0.0	100	0.00
68	T	1,2-Dibromoethane	-1.0000	0.3394	0.0	100	0.00
69	T	1-Chlorohexane	-1.0000	0.3164	0.0	100	0.00
70	P	Chlorobenzene	-1.0000	0.3328	0.0	100	0.00
71	T	1,1,1,2-Tetrachloroethane	-1.0000	0.3096	0.0	100	0.00
72	C	Ethylbenzene	-1.0000	0.2934	0.0	100	-0.01
73	T	m-,p-Xylene	-1.0000	0.6799	0.0	100	0.00
74	T	o-Xylene	-1.0000	0.3047	0.0	100	0.00
75	T	Styrene	-1.0000	0.3062	0.0	100	0.00
76	P	Bromoform	-1.0000	0.2681	0.0	100	0.00
77	T	Isopropylbenzene	-1.0000	0.3287	0.0	100	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
79	P	1,1,2,2-Tetrachloroethane	-1.0000	0.2853	0.0	100	0.00
80	S	p-Bromofluorobenzene	-1.0000	0.0253	0.0	100	0.00
81	T	1,2,3-Trichloropropane	-1.0000	0.1103	0.0	100	0.00
82	T	trans-1,4-Dichloro-2-Butene	-1.0000	0.0000	0.0	0	-16.19#
83	T	n-Propylbenzene	-1.0000	0.3357	0.0	100	0.00
84	T	Bromobenzene	0.3000	0.2977	0.8	100	0.00
85	T	1,3,5-Trimethylbenzene	-1.0000	0.3196	0.0	100	0.00
86	T	2-Chlorotoluene	-1.0000	0.3415	0.0	100	0.00
87	T	4-Chlorotoluene	-1.0000	0.3499	0.0	100	0.00
88	T	a-Methylstyrene	-1.0000	0.2534	0.0	100	0.00
89	T	tert-Butylbenzene	-1.0000	0.2606	0.0	100	0.00
90	T	1,2,4-Trimethylbenzene	-1.0000	0.3069	0.0	100	0.00
91	T	sec-Butylbenzene	-1.0000	0.3436	0.0	100	0.00
92	T	p-Isopropyltoluene	-1.0000	0.3192	0.0	100	0.00
93	T	1,3-Dichlorobenzene	-1.0000	0.3257	0.0	100	0.00
94	T	1,4-Dichlorobenzene	0.3000	0.3299	-10.0	100	0.00
95	T	n-Butylbenzene	-1.0000	0.3323	0.0	100	0.00
96	T	1,2-Dichlorobenzene	0.3000	0.3472	-15.7	100	0.00
97	T	1,2-Dibromo-3-Chloropropane	-1.0000	0.0000	0.0	0	-19.13#
98	T	1,2,4-Trichlorobenzene	-1.0000	0.3418	0.0	100	0.00
99	T	Hexachlorobutadiene	-1.0000	0.3297	0.0	100	0.00
100	T	Naphthalene	-1.0000	0.3306	0.0	100	0.00
101	T	1,2,3-Trichlorobenzene	0.3000	0.3161	-5.4	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412220.D 8260WT.M Sat May 14 18:03:34 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412221.D Vial: 4
Acq On : 13 May 2016 13:58 Operator: TMB
Sample : WG568561-03 0.4ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:20:36 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	801173	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	557498	25.00 ug/L	0.01
78) 1,4-Dichlorobenzene-d4	17.60	152	293108	25.00 ug/L	0.01

System Monitoring Compounds					
37) Dibromofluoromethane	0.00	111	0	0.0000 ug/L	
Spiked Amount 25.000	Range 86 - 118		Recovery =	0.00%#	
43) 1,2-Dichloroethane-d4	0.00	65	0	0.0000 ug/L	
Spiked Amount 25.000	Range 80 - 120		Recovery =	0.00%#	
58) Toluene-d8	0.00	98	0	0.0000 ug/L	
Spiked Amount 25.000	Range 88 - 110		Recovery =	0.00%#	
80) p-Bromofluorobenzene	16.08	95	343	0.0286 ug/L	0.01
Spiked Amount 25.000	Range 86 - 115		Recovery =	0.12%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	5478	0.3262 ug/L	79
3) Chloromethane	3.63	50	3930	0.2484 ug/L	85
4) Vinyl Chloride	3.84	62	3512	0.3288 ug/L	97
5) 1,3-Butadiene	3.90	54	2479	0.2375 ug/L #	75
6) Bromomethane	4.73	94	1343	0.9497 ug/L	85
7) Chloroethane	4.90	64	2219	0.3174 ug/L #	43
8) Trichlorofluoromethane	5.36	101	6974	0.3289 ug/L #	93
10) Isoprene	5.94	67	4288	0.3952 ug/L #	62
12) 1,1,2-Trichloro-1,2,2-Trif	6.17	101	3168	0.3263 ug/L #	40
14) 1,1-Dichloroethene	6.47	61	5431	0.2882 ug/L	92
16) Dimethyl Sulfide	6.74	62	1708	0.2453 ug/L	64
19) Methylene Chloride	7.24	84	3785	0.4239 ug/L #	43
20) Carbon Disulfide	7.29	76	12437	0.4565 ug/L	97
22) Methyl Tert Butyl Ether	7.47	73	6580	0.3336 ug/L #	76
23) trans-1,2-Dichloroethene	7.71	61	4990	0.2928 ug/L	96
24) n-Hexane	7.79	57	3110	0.2011 ug/L	99
26) Vinyl Acetate	8.31	43	2753	0.1566 ug/L	100
27) 1,1-Dichloroethane	8.32	63	5655	0.2774 ug/L #	92
31) 2,2-Dichloropropane	9.10	77	6035	0.3360 ug/L	82
32) cis-1,2-Dichloroethene	9.16	96	3837	0.3762 ug/L	74
33) Chloroform	9.37	83	7267	0.3842 ug/L	100
34) 1-Bromopropane	9.50	122	234	0.1456 ug/L #	1
35) Bromochloromethane	9.60	130	1674	0.3137 ug/L	62
36) Tetrahydrofuran	9.62	42	447	0.2146 ug/L #	41
38) 1,1,1-Trichloroethane	9.91	97	6304	0.3305 ug/L	90
39) Cyclohexane	9.94	56	4658	0.2238 ug/L #	67
40) 1,1-Dichloropropene	10.11	75	5172	0.3768 ug/L	80
42) Carbon Tetrachloride	10.25	117	5692	0.3172 ug/L	95
45) 1,2-Dichloroethane	10.43	62	4443	0.2796 ug/L #	63
46) Benzene	10.46	78	14456	0.4134 ug/L #	77
47) Trichloroethene	11.21	130	3659	0.3626 ug/L	98
48) Methylcyclohexane	11.29	83	5025	0.3562 ug/L	84
49) 1,2-Dichloropropane	11.42	63	3060	0.3007 ug/L	98
50) Bromodichloromethane	11.73	83	4734	0.3343 ug/L #	93
52) Dibromomethane	11.82	93	1805	0.3767 ug/L	84
55) cis-1,3-Dichloropropene	12.36	75	5008	0.3514 ug/L	90
56) Dimethyl Disulfide	12.63	79	1921	0.2573 ug/L	85
59) Toluene	12.78	91	14742	0.4416 ug/L	98
60) Ethyl Methacrylate	12.90	69	2062	0.3292 ug/L	66
62) trans-1,3-Dichloropropene	12.96	75	3693	0.3171 ug/L	100

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412221.D Vial: 4
Acq On : 13 May 2016 13:58 Operator: TMB
Sample : WG568561-03 0.4ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:20:36 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
63)	1,1,2-Trichloroethane	13.18	97	2326	0.4391 ug/L		70
65)	1,3-Dichloropropane	13.49	76	3163	0.3219 ug/L		88
66)	Tetrachloroethene	13.61	164	3164	0.4185 ug/L		93
67)	Dibromochloromethane	13.87	129	3113	0.3727 ug/L		86
68)	1,2-Dibromoethane	14.12	107	2076	0.3760 ug/L		90
69)	1-Chlorohexane	14.21	91	4746	0.4174 ug/L		80
70)	Chlorobenzene	14.62	112	9180	0.3960 ug/L		98
71)	1,1,1,2-Tetrachloroethane	14.66	131	3501	0.3876 ug/L		95
72)	Ethylbenzene	14.65	106	5348	0.4066 ug/L		84
73)	m-,p-Xylene	14.75	106	13269	0.8757 ug/L		89
74)	o-Xylene	15.31	106	6019	0.3943 ug/L		97
75)	Styrene	15.35	104	9370	0.3849 ug/L		95
76)	Bromoform	15.83	173	1198	0.2500 ug/L	#	76
77)	Isopropylbenzene	15.73	105	16528	0.4270 ug/L		92
79)	1,1,2,2-Tetrachloroethane	15.96	83	1999	0.3419 ug/L	#	83
83)	n-Propylbenzene	16.25	91	19410	0.4297 ug/L		95
84)	Bromobenzene	16.37	156	3319	0.3524 ug/L		75
85)	1,3,5-Trimethylbenzene	16.42	105	13794	0.4127 ug/L		95
86)	2-Chlorotoluene	16.51	91	13678	0.4729 ug/L		90
87)	4-Chlorotoluene	16.56	91	12572	0.4219 ug/L		98
88)	a-Methylstyrene	16.84	118	5739	0.3485 ug/L		96
89)	tert-Butylbenzene	16.90	134	2254	0.3403 ug/L		55
90)	1,2,4-Trimethylbenzene	16.95	105	13859	0.4051 ug/L		90
91)	sec-Butylbenzene	17.17	105	17612	0.4382 ug/L		90
92)	p-Isopropyltoluene	17.32	119	13616	0.3968 ug/L		90
93)	1,3-Dichlorobenzene	17.52	146	6961	0.3724 ug/L		86
94)	1,4-Dichlorobenzene	17.64	146	7149	0.3832 ug/L	#	1
95)	n-Butylbenzene	17.85	91	15223	0.4664 ug/L		91
96)	1,2-Dichlorobenzene	18.14	146	6129	0.3742 ug/L		92
98)	1,2,4-Trichlorobenzene	20.28	180	5035	0.3969 ug/L		97
99)	Hexachlorobutadiene	20.44	225	2420	0.4122 ug/L		83
100)	Naphthalene	20.65	128	7474	0.3874 ug/L	#	84
101)	1,2,3-Trichlorobenzene	20.96	180	4378	0.4056 ug/L	#	78

(#) = qualifier out of range (m) = manual integration
8M412221.D 8260WT.M Fri May 13 14:20:39 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412221.D Vial: 4
Acq On : 13 May 2016 13:58 Operator: TMB
Sample : WG568561-03 0.4ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:20 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412221.D Vial: 4
Acq On : 13 May 2016 13:58 Operator: TMB
Sample : WG568561-03 0.4ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:03:01 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	0.4000	0.4078	-2.0	100	0.00
3	P	Chloromethane	-1.0000	0.5012	0.0	100	0.00
4	C	Vinyl Chloride	0.4000	0.3775	5.6	100	0.00
5	T	1,3-Butadiene	-1.0000	0.3178	0.0	100	0.02
6	T	Bromomethane	-1.0000	1.0253	0.0	100	0.00
7	T	Chloroethane	0.4000	0.4335	-8.4	100	0.00
8	T	Trichlorofluoromethane	0.4000	0.4029	-0.7	100	-0.01
9	T	Diethyl ether	-1.0000	0.0000	0.0	0	-5.90#
10	T	Isoprene	0.4000	0.4079	-2.0	100	0.00
11	T	Acrolein	-1.0000	0.0000	0.0	0	-6.13#
12	T	1,1,2-Trichloro-1,2,2-Trifl	-1.0000	0.3894	0.0	100	0.02
13	T	Acetone	-1.0000	0.0000	0.0	0	-6.25#
14	C	1,1-Dichloroethene	0.4000	0.4192	-4.8	100	0.00
15	T	Tert-Butyl Alcohol	-1.0000	0.0000	0.0	0	-6.58#
16	T	Dimethyl Sulfide	0.4000	0.3687	7.8	100	0.00
17	T	Iodomethane	-1.0000	0.0000	0.0	0	-6.98#
18	T	Methyl acetate	-1.0000	0.2179	0.0	100	0.00
19	T	Methylene Chloride	-1.0000	0.4591	0.0	100	0.00
20	T	Carbon Disulfide	-1.0000	0.4587	0.0	100	0.00
21	T	Acrylonitrile	-1.0000	0.0000	0.0	0	-7.43#
22	T	Methyl Tert Butyl Ether	-1.0000	0.3688	0.0	100	0.00
23	T	trans-1,2-Dichloroethene	0.4000	0.4193	-4.8	100	0.00
24	T	n-Hexane	-1.0000	0.3519	0.0	100	0.00
25	T	Diisopropyl ether	-1.0000	0.0000	0.0	0	-8.13#
26	T	Vinyl Acetate	-1.0000	0.3241	0.0	100	0.02
27	P	1,1-Dichloroethane	0.4000	0.3826	4.4	100	0.00
28	T	Ethyl-Tert-Butyl ether	-1.0000	0.0000	0.0	0	-8.71#
29	T	2-Butanone	-1.0000	0.0000	0.0	0	-8.88#
30	T	Propionitrile	-1.0000	0.0000	0.0	0	-8.98#
31	T	2,2-Dichloropropane	0.4000	0.3874	3.1	100	0.00
32	T	cis-1,2-Dichloroethene	0.4000	0.4172	-4.3	100	-0.01
33	C	Chloroform	0.4000	0.4164	-4.1	100	0.00
34		1-Bromopropane	-1.0000	0.1776	0.0	100	-0.01
35	T	Bromochloromethane	0.4000	0.3550	11.3	100	0.00
36	T	Tetrahydrofuran	-1.0000	0.5125	0.0	100	-0.01
37	S	Dibromofluoromethane	-1.0000	0.0000	0.0	0	-9.67#
38	T	1,1,1-Trichloroethane	0.4000	0.3803	4.9	100	0.00
39	T	Cyclohexane	-1.0000	0.4058	0.0	100	0.00
40	T	1,1-Dichloropropene	-1.0000	0.4093	0.0	100	0.00
41	T	Tert-Amyl-Methyl ether	-1.0000	0.0000	0.0	0	-10.21#
42	T	Carbon Tetrachloride	0.4000	0.3867	3.3	100	0.00
43	S	1,2-Dichloroethane-d4	-1.0000	0.0000	0.0	0	-10.31#
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	0.4000	0.3955	1.1	100	0.00
46	T	Benzene	0.4000	0.4393	-9.8	100	0.00
47	T	Trichloroethene	0.4000	0.4169	-4.2	100	0.00
48	T	Methylcyclohexane	-1.0000	0.3768	0.0	100	-0.01
49	C	1,2-Dichloropropane	0.4000	0.4177	-4.4	100	0.00
50	T	Bromodichloromethane	0.4000	0.3676	8.1	100	0.00
51	T	1,4-Dioxane	-1.0000	0.0000	0.0	0	-11.71#
52	T	Dibromomethane	0.4000	0.4162	-4.1	100	0.00
53	T	2-Chloroethyl Vinyl Ether	-1.0000	0.1401	0.0	100	0.00
54	T	4-Methyl-2-Pentanone	-1.0000	0.0000	0.0	0	-12.05#

(#) = Out of Range

8M412221.D 8260WT.M Sat May 14 18:03:54 2016 Page 1

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412221.D Vial: 4
Acq On : 13 May 2016 13:58 Operator: TMB
Sample : WG568561-03 0.4ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:03:01 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	0.4000	0.3821	4.5	100	0.00
56	T	Dimethyl Disulfide	-1.0000	0.2746	0.0	100	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	-1.0000	0.0000	0.0	0	-12.68#
59	C	Toluene	0.4000	0.4259	-6.5	100	0.00
60	T	Ethyl Methacrylate	-1.0000	0.3102	0.0	100	0.02
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	-1.0000	0.3184	0.0	100	0.00
63	T	1,1,2-Trichloroethane	0.4000	0.4259	-6.5	100	0.00
64	T	2-Hexanone	-1.0000	0.0000	0.0	0	-13.12#
65	T	1,3-Dichloropropane	0.4000	0.3253	18.7	100	0.00
66	T	Tetrachloroethene	0.4000	0.4267	-6.7	100	0.00
67	T	Dibromochloromethane	0.4000	0.3904	2.4	100	0.00
68	T	1,2-Dibromoethane	0.4000	0.3814	4.6	100	0.00
69	T	1-Chlorohexane	0.4000	0.4017	-0.4	100	0.00
70	P	Chlorobenzene	0.4000	0.3933	1.7	100	0.00
71	T	1,1,1,2-Tetrachloroethane	0.4000	0.3872	3.2	100	0.00
72	C	Ethylbenzene	0.4000	0.4188	-4.7	100	0.00
73	T	m-,p-Xylene	0.8000	0.8859	-10.7	100	0.00
74	T	o-Xylene	-1.0000	0.4110	0.0	100	0.00
75	T	Styrene	0.4000	0.3901	2.5	100	0.00
76	P	Bromoform	-1.0000	0.2533	0.0	100	0.00
77	T	Isopropylbenzene	0.4000	0.4249	-6.2	100	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
79	P	1,1,2,2-Tetrachloroethane	0.4000	0.3398	15.1	100	0.00
80	S	p-Bromofluorobenzene	-1.0000	0.0283	0.0	100	0.00
81	T	1,2,3-Trichloropropane	-1.0000	0.0000	0.0	0	-16.14#
82	T	trans-1,4-Dichloro-2-Butene	-1.0000	0.0000	0.0	0	-16.19#
83	T	n-Propylbenzene	0.4000	0.4245	-6.1	100	0.00
84	T	Bromobenzene	0.4000	0.3529	11.8	100	0.00
85	T	1,3,5-Trimethylbenzene	0.4000	0.4141	-3.5	100	0.00
86	T	2-Chlorotoluene	0.4000	0.4284	-7.1	100	0.00
87	T	4-Chlorotoluene	0.4000	0.4359	-9.0	100	0.00
88	T	a-Methylstyrene	-1.0000	0.3631	0.0	100	0.00
89	T	tert-Butylbenzene	-1.0000	0.3459	0.0	100	0.00
90	T	1,2,4-Trimethylbenzene	0.4000	0.4046	-1.2	100	0.00
91	T	sec-Butylbenzene	-1.0000	0.4494	0.0	100	0.00
92	T	p-Isopropyltoluene	-1.0000	0.4092	0.0	100	0.00
93	T	1,3-Dichlorobenzene	0.4000	0.3844	3.9	100	0.00
94	T	1,4-Dichlorobenzene	0.4000	0.3915	2.1	100	0.00
95	T	n-Butylbenzene	-1.0000	0.4628	0.0	100	0.00
96	T	1,2-Dichlorobenzene	0.4000	0.3805	4.9	100	0.00
97	T	1,2-Dibromo-3-Chloropropane	-1.0000	0.0000	0.0	0	-19.13#
98	T	1,2,4-Trichlorobenzene	0.4000	0.4182	-4.5	100	0.00
99	T	Hexachlorobutadiene	0.4000	0.4302	-7.5	100	0.00
100	T	Naphthalene	0.4000	0.4178	-4.5	100	0.00
101	T	1,2,3-Trichlorobenzene	0.4000	0.4310	-7.7	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412221.D 8260WT.M Sat May 14 18:03:56 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412222.D Vial: 5
Acq On : 13 May 2016 14:27 Operator: TMB
Sample : WG568561-04 1ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:49:45 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	795577	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	547129	25.00 ug/L	0.01
78) 1,4-Dichlorobenzene-d4	17.60	152	284579	25.00 ug/L	0.01
System Monitoring Compounds					
37) Dibromofluoromethane	9.65	111	3962	0.3966 ug/L	-0.02
Spiked Amount 25.000	Range 86 - 118		Recovery	= 1.60%#	
43) 1,2-Dichloroethane-d4	10.31	65	3835	0.3356 ug/L	0.01
Spiked Amount 25.000	Range 80 - 120		Recovery	= 1.36%#	
58) Toluene-d8	12.68	98	14023	0.4957 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery	= 2.00%#	
80) p-Bromofluorobenzene	16.08	95	6483	0.5565 ug/L	0.01
Spiked Amount 25.000	Range 86 - 115		Recovery	= 2.24%#	
Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	12402	0.7438 ug/L	97
3) Chloromethane	3.62	50	8415	0.5357 ug/L #	75
4) Vinyl Chloride	3.84	62	9412	0.8873 ug/L	89
5) 1,3-Butadiene	3.90	54	6974	0.6729 ug/L	81
6) Bromomethane	4.73	94	4114	1.3123 ug/L	92
7) Chloroethane	4.90	64	4983	0.7178 ug/L #	80
8) Trichlorofluoromethane	5.38	101	17133	0.8136 ug/L #	96
9) Diethyl ether	5.90	59	20707	3.2140 ug/L	84
10) Isoprene	5.94	67	11176	1.0373 ug/L	76
11) Acrolein	6.14	56	246	0.4300 ug/L #	15
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	8135	0.8438 ug/L	96
13) Acetone	6.24	43	413	0.2115 ug/L #	48
14) 1,1-Dichloroethene	6.46	61	12947	0.6919 ug/L	91
15) Tert-Butyl Alcohol	6.59	59	3316	7.3747 ug/L #	55
16) Dimethyl Sulfide	6.72	62	4505	0.6515 ug/L	79
18) Methyl acetate	7.00	43	2119	0.3630 ug/L #	66
19) Methylene Chloride	7.24	84	8723	0.9837 ug/L #	43
20) Carbon Disulfide	7.29	76	28246	1.0441 ug/L	100
21) Acrylonitrile	7.43	53	2164	0.8620 ug/L	76
22) Methyl Tert Butyl Ether	7.47	73	16998	0.8678 ug/L	84
23) trans-1,2-Dichloroethene	7.70	61	12571	0.7427 ug/L	100
24) n-Hexane	7.78	57	9382	0.6109 ug/L	82
25) Diisopropyl ether	8.12	45	95400	2.1939 ug/L #	91
26) Vinyl Acetate	8.30	43	6750	0.3867 ug/L #	89
27) 1,1-Dichloroethane	8.32	63	14721	0.7271 ug/L	98
28) Ethyl-Tert-Butyl ether	8.70	59	97972	3.2102 ug/L	88
29) 2-Butanone	8.88	43	380	0.1215 ug/L #	57
30) Propionitrile	8.99	54	1593	2.0138 ug/L #	56
31) 2,2-Dichloropropane	9.09	77	14691	0.8236 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	9482	0.9362 ug/L	72
33) Chloroform	9.37	83	18494	0.9845 ug/L	94
34) 1-Bromopropane	9.50	122	1092	0.6844 ug/L	80
35) Bromochloromethane	9.61	130	5234	0.9877 ug/L #	55
36) Tetrahydrofuran	9.63	42	5415	2.6185 ug/L #	68
38) 1,1,1-Trichloroethane	9.91	97	17074	0.9016 ug/L	92
39) Cyclohexane	9.94	56	11865	0.5740 ug/L #	67
40) 1,1-Dichloropropene	10.11	75	13172	0.9663 ug/L	87
41) Tert-Amyl-Methyl ether	10.21	73	90255	4.6173 ug/L #	88
42) Carbon Tetrachloride	10.25	117	14515	0.8146 ug/L	95
45) 1,2-Dichloroethane	10.42	62	10675	0.6766 ug/L	85

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412222.D Vial: 5
Acq On : 13 May 2016 14:27 Operator: TMB
Sample : WG568561-04 1ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:49:45 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
46)	Benzene	10.47	78	34504	0.9938 ug/L		84
47)	Trichloroethene	11.21	130	8752	0.8735 ug/L		93
48)	Methylcyclohexane	11.30	83	13513	0.9646 ug/L		79
49)	1,2-Dichloropropane	11.43	63	7057	0.6982 ug/L	#	67
50)	Bromodichloromethane	11.73	83	12147	0.8638 ug/L		96
52)	Dibromomethane	11.81	93	4130	0.8681 ug/L		95
53)	2-Chloroethyl Vinyl Ether	12.03	63	2694	0.6366 ug/L		86
54)	4-Methyl-2-Pentanone	12.06	58	848	0.3671 ug/L	#	41
55)	cis-1,3-Dichloropropene	12.36	75	12521	0.8847 ug/L		97
56)	Dimethyl Disulfide	12.63	79	5799	0.7821 ug/L		96
59)	Toluene	12.79	91	37114	1.1328 ug/L		95
60)	Ethyl Methacrylate	12.88	69	5490	0.8930 ug/L		69
62)	trans-1,3-Dichloropropene	12.96	75	10328	0.9035 ug/L		92
63)	1,1,2-Trichloroethane	13.18	97	5274	1.0144 ug/L		93
64)	2-Hexanone	13.10	58	397	0.2127 ug/L	#	1
65)	1,3-Dichloropropane	13.49	76	9673	1.0032 ug/L		73
66)	Tetrachloroethene	13.61	164	7373	0.9937 ug/L		95
67)	Dibromochloromethane	13.87	129	7358	0.8977 ug/L		95
68)	1,2-Dibromoethane	14.13	107	5123	0.9454 ug/L		95
69)	1-Chlorohexane	14.21	91	12790	1.1461 ug/L		79
70)	Chlorobenzene	14.62	112	25662	1.1281 ug/L		89
71)	1,1,1,2-Tetrachloroethane	14.66	131	8722	0.9838 ug/L		94
72)	Ethylbenzene	14.65	106	13009	1.0078 ug/L		87
73)	m-,p-Xylene	14.75	106	30465	2.0486 ug/L		84
74)	o-Xylene	15.31	106	15133	1.0101 ug/L		84
75)	Styrene	15.35	104	23191	0.9707 ug/L		91
76)	Bromoform	15.84	173	3875	0.8240 ug/L		96
77)	Isopropylbenzene	15.74	105	41382	1.0895 ug/L		95
79)	1,1,2,2-Tetrachloroethane	15.95	83	5604	0.9873 ug/L	#	93
81)	1,2,3-Trichloropropane	16.15	110	1355	0.7428 ug/L	#	1
82)	trans-1,4-Dichloro-2-Buten	16.19	53	962	0.3479 ug/L	#	1
83)	n-Propylbenzene	16.25	91	48249	1.1002 ug/L		99
84)	Bromobenzene	16.37	156	9598	1.0497 ug/L		95
85)	1,3,5-Trimethylbenzene	16.42	105	33355	1.0278 ug/L		96
86)	2-Chlorotoluene	16.51	91	33408	1.1895 ug/L		94
87)	4-Chlorotoluene	16.56	91	29548	1.0213 ug/L		95
88)	a-Methylstyrene	16.84	118	14414	0.9014 ug/L		91
89)	tert-Butylbenzene	16.89	134	6686	1.0398 ug/L		86
90)	1,2,4-Trimethylbenzene	16.95	105	36066	1.0857 ug/L		96
91)	sec-Butylbenzene	17.16	105	42128	1.0796 ug/L		96
92)	p-Isopropyltoluene	17.32	119	35424	1.0633 ug/L		94
93)	1,3-Dichlorobenzene	17.52	146	18409	1.0144 ug/L		93
94)	1,4-Dichlorobenzene	17.64	146	19031	1.0506 ug/L	#	62
95)	n-Butylbenzene	17.85	91	35772	1.1288 ug/L		96
96)	1,2-Dichlorobenzene	18.14	146	16104	1.0126 ug/L		92
97)	1,2-Dibromo-3-Chloropropan	19.14	75	1262	1.0152 ug/L	#	53
98)	1,2,4-Trichlorobenzene	20.28	180	11494	0.9331 ug/L		98
99)	Hexachlorobutadiene	20.43	225	5421	0.9510 ug/L		98
100)	Naphthalene	20.65	128	16581	0.8853 ug/L		100
101)	1,2,3-Trichlorobenzene	20.96	180	10306	0.9835 ug/L		94

(#) = qualifier out of range (m) = manual integration
8M412222.D 8260WT.M Fri May 13 14:49:48 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412222.D Vial: 5
Acq On : 13 May 2016 14:27 Operator: TMB
Sample : WG568561-04 1ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:49 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration



8M412222.D 8260WT.M Fri May 13 14:49:48 2016 Page 3

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412222.D Vial: 5
Acq On : 13 May 2016 14:27 Operator: TMB
Sample : WG568561-04 1ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:05:31 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	1.0000	0.9298	7.0	100	0.00
3	P	Chloromethane	1.0000	1.0808	-8.1	100	0.00
4	C	Vinyl Chloride	1.0000	1.0189	-1.9	100	0.00
5	T	1,3-Butadiene	-1.0000	0.9002	0.0	100	0.02
6	T	Bromomethane	1.0000	1.2082	-20.8	100	0.00
7	T	Chloroethane	1.0000	0.9803	2.0	100	0.00
8	T	Trichlorofluoromethane	1.0000	0.9968	0.3	100	0.00
9	T	Diethyl ether	5.0000	5.0462	-0.9	100	0.00
10	T	Isoprene	1.0000	1.0707	-7.1	100	0.00
11	T	Acrolein	-1.0000	0.6696	0.0	100	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	1.0000	1.0069	-0.7	100	0.00
13	T	Acetone	-1.0000	0.4614	0.0	100	0.00
14	C	1,1-Dichloroethene	1.0000	1.0064	-0.6	100	0.00
15	T	Tert-Butyl Alcohol	-1.0000	11.5604	0.0	0	0.02
16	T	Dimethyl Sulfide	1.0000	0.9794	2.1	100	-0.01
17	T	Iodomethane	-1.0000	0.0000	0.0	0	-6.98#
18	T	Methyl acetate	-1.0000	0.8758	0.0	100	0.00
19	T	Methylene Chloride	1.0000	1.0655	-6.5	100	0.00
20	T	Carbon Disulfide	1.0000	1.0492	-4.9	100	0.00
21	T	Acrylonitrile	2.5000	1.8633	25.5#	100	0.00
22	T	Methyl Tert Butyl Ether	1.0000	0.9593	4.1	100	0.00
23	T	trans-1,2-Dichloroethene	1.0000	1.0636	-6.4	100	0.00
24	T	n-Hexane	-1.0000	1.0690	0.0	100	0.00
25	T	Diisopropyl ether	5.0000	5.1784	-3.6	100	0.00
26	T	Vinyl Acetate	-1.0000	0.8002	0.0	100	0.00
27	P	1,1-Dichloroethane	1.0000	1.0029	-0.3	100	0.00
28	T	Ethyl-Tert-Butyl ether	5.0000	4.9635	0.7	100	0.00
29	T	2-Butanone	-1.0000	0.2889	0.0	100	0.00
30	T	Propionitrile	5.0000	4.1566	16.9	100	0.00
31	T	2,2-Dichloropropane	1.0000	0.9498	5.0	100	0.00
32	T	cis-1,2-Dichloroethene	1.0000	1.0383	-3.8	100	0.00
33	C	Chloroform	1.0000	1.0671	-6.7	100	0.00
34		1-Bromopropane	1.0000	0.8348	16.5	100	-0.01
35	T	Bromochloromethane	1.0000	1.1177	-11.8	100	0.00
36	T	Tetrahydrofuran	5.0000	6.2516	-25.0#	100	0.00
37	S	Dibromofluoromethane	0.5000	0.4635	7.3	100	-0.01
38	T	1,1,1-Trichloroethane	1.0000	1.0372	-3.7	100	0.00
39	T	Cyclohexane	1.0000	1.0411	-4.1	100	0.00
40	T	1,1-Dichloropropene	1.0000	1.0498	-5.0	100	0.00
41	T	Tert-Amyl-Methyl ether	5.0000	5.0344	-0.7	100	0.00
42	T	Carbon Tetrachloride	1.0000	0.9930	0.7	100	0.00
43	S	1,2-Dichloroethane-d4	0.5000	0.4184	16.3	100	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	1.0000	0.9570	4.3	100	0.00
46	T	Benzene	1.0000	1.0558	-5.6	100	0.00
47	T	Trichloroethene	1.0000	1.0041	-0.4	100	0.00
48	T	Methylcyclohexane	-1.0000	1.0205	0.0	100	0.00
49	C	1,2-Dichloropropane	1.0000	0.9700	3.0	100	0.00
50	T	Bromodichloromethane	1.0000	0.9500	5.0	100	0.00
51	T	1,4-Dioxane	-1.0000	0.0000	0.0	0	-11.71#
52	T	Dibromomethane	1.0000	0.9590	4.1	100	0.00
53	T	2-Chloroethyl Vinyl Ether	-1.0000	0.8616	0.0	100	0.00
54	T	4-Methyl-2-Pentanone	-1.0000	0.6639	0.0	100	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412222.D Vial: 5
Acq On : 13 May 2016 14:27 Operator: TMB
Sample : WG568561-04 1ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:05:31 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	1.0000	0.9621	3.8	100	0.00
56	T	Dimethyl Disulfide	-1.0000	0.8348	0.0	100	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	0.5000	0.4897	2.1	100	0.00
59	C	Toluene	1.0000	1.0927	-9.3	100	0.00
60	T	Ethyl Methacrylate	1.0000	0.8416	15.8	100	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	1.0000	0.9073	9.3	100	0.00
63	T	1,1,2-Trichloroethane	1.0000	0.9840	1.6	100	0.00
64	T	2-Hexanone	-1.0000	0.3525	0.0	100	-0.01
65	T	1,3-Dichloropropane	1.0000	1.0137	-1.4	100	0.00
66	T	Tetrachloroethene	1.0000	1.0132	-1.3	100	0.00
67	T	Dibromochloromethane	1.0000	0.9402	6.0	100	0.00
68	T	1,2-Dibromoethane	1.0000	0.9590	4.1	100	0.00
69	T	1-Chlorohexane	1.0000	1.1031	-10.3	100	0.00
70	P	Chlorobenzene	1.0000	1.1204	-12.0	100	0.00
71	T	1,1,1,2-Tetrachloroethane	1.0000	0.9829	1.7	100	0.00
72	C	Ethylbenzene	1.0000	1.0381	-3.8	100	0.00
73	T	m-,p-Xylene	2.0000	2.0726	-3.6	100	0.00
74	T	o-Xylene	1.0000	1.0530	-5.3	100	0.00
75	T	Styrene	1.0000	0.9839	1.6	100	0.00
76	P	Bromoform	1.0000	0.8348	16.5	100	0.00
77	T	Isopropylbenzene	1.0000	1.0840	-8.4	100	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
79	P	1,1,2,2-Tetrachloroethane	1.0000	0.9810	1.9	100	0.00
80	S	p-Bromofluorobenzene	-1.0000	0.5502	0.0	100	0.00
81	T	1,2,3-Trichloropropane	1.0000	0.7703	23.0	100	0.02
82	T	trans-1,4-Dichloro-2-Butene	1.0000	1.1443	-14.4	100	0.00
83	T	n-Propylbenzene	1.0000	1.0868	-8.7	100	0.00
84	T	Bromobenzene	1.0000	1.0510	-5.1	100	0.00
85	T	1,3,5-Trimethylbenzene	1.0000	1.0314	-3.1	100	0.00
86	T	2-Chlorotoluene	1.0000	1.0778	-7.8	100	0.00
87	T	4-Chlorotoluene	1.0000	1.0553	-5.5	100	0.00
88	T	a-Methylstyrene	-1.0000	0.9394	0.0	100	0.00
89	T	tert-Butylbenzene	1.0000	1.0567	-5.7	100	0.00
90	T	1,2,4-Trimethylbenzene	1.0000	1.0845	-8.5	100	0.00
91	T	sec-Butylbenzene	1.0000	1.1073	-10.7	100	0.00
92	T	p-Isopropyltoluene	1.0000	1.0966	-9.7	100	0.00
93	T	1,3-Dichlorobenzene	1.0000	1.0469	-4.7	100	0.00
94	T	1,4-Dichlorobenzene	1.0000	1.0735	-7.3	100	0.00
95	T	n-Butylbenzene	1.0000	1.1201	-12.0	100	0.00
96	T	1,2-Dichlorobenzene	1.0000	1.0297	-3.0	100	0.00
97	T	1,2-Dibromo-3-Chloropropane	-1.0000	1.1006	0.0	100	0.00
98	T	1,2,4-Trichlorobenzene	1.0000	0.9832	1.7	100	0.00
99	T	Hexachlorobutadiene	1.0000	0.9925	0.7	100	0.00
100	T	Naphthalene	1.0000	0.9547	4.5	100	0.00
101	T	1,2,3-Trichlorobenzene	1.0000	1.0450	-4.5	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412222.D 8260WT.M Sat May 14 18:05:50 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412223.D Vial: 6
Acq On : 13 May 2016 14:56 Operator: TMB
Sample : WG568561-05 2ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:18:34 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	10.71	96	782460	25.00	ug/L	0.00
57) Chlorobenzene-d5	14.57	117	543252	25.00	ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	286450	25.00	ug/L	0.01

System Monitoring Compounds						
37) Dibromofluoromethane	9.66	111	8161	0.8307	ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery =	3.32%#		
43) 1,2-Dichloroethane-d4	10.30	65	9681	0.8613	ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery =	3.44%#		
58) Toluene-d8	12.68	98	28135	1.0016	ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery =	4.00%#		
80) p-Bromofluorobenzene	16.08	95	12079	1.0302	ug/L	0.01
Spiked Amount 25.000	Range 86 - 115		Recovery =	4.12%#		

Target Compounds						Qvalue
2) Dichlorodifluoromethane	3.16	85	24870	1.5166	ug/L	98
3) Chloromethane	3.62	50	16106	1.0425	ug/L	93
4) Vinyl Chloride	3.85	62	18562	1.7792	ug/L	94
5) 1,3-Butadiene	3.90	54	13617	1.3360	ug/L	92
6) Bromomethane	4.73	94	8329	1.8804	ug/L	94
7) Chloroethane	4.90	64	9562	1.4006	ug/L	86
8) Trichlorofluoromethane	5.39	101	34970	1.6885	ug/L	97
9) Diethyl ether	5.90	59	107986	17.0420	ug/L	79
10) Isoprene	5.93	67	20936	1.9757	ug/L	73
11) Acrolein	6.14	56	4864	8.6444	ug/L	95
12) 1,1,2-Trichloro-1,2,2-Trif	6.17	101	16832	1.7751	ug/L	99
13) Acetone	6.23	43	1519	0.7908	ug/L #	48
14) 1,1-Dichloroethene	6.47	61	25394	1.3798	ug/L	89
15) Tert-Butyl Alcohol	6.58	59	14891	33.6725	ug/L	85
16) Dimethyl Sulfide	6.73	62	9489	1.3953	ug/L	79
17) Iodomethane	6.98	142	1211	1.3472	ug/L #	60
18) Methyl acetate	7.01	43	5064	0.8821	ug/L #	79
19) Methylene Chloride	7.25	84	16579	1.9010	ug/L	52
20) Carbon Disulfide	7.29	76	57693	2.1683	ug/L	98
21) Acrylonitrile	7.43	53	13848	5.6089	ug/L	93
22) Methyl Tert Butyl Ether	7.47	73	33991	1.7644	ug/L	84
23) trans-1,2-Dichloroethene	7.70	61	22786	1.3688	ug/L	88
24) n-Hexane	7.79	57	18282	1.2105	ug/L	84
25) Diisopropyl ether	8.12	45	475271	11.1128	ug/L #	92
26) Vinyl Acetate	8.29	43	15998	0.9318	ug/L	93
27) 1,1-Dichloroethane	8.32	63	30419	1.5277	ug/L	97
28) Ethyl-Tert-Butyl ether	8.70	59	505010	16.8250	ug/L	89
29) 2-Butanone	8.90	43	2250	0.7313	ug/L #	57
30) Propionitrile	8.99	54	9373	12.0479	ug/L	86
31) 2,2-Dichloropropane	9.10	77	32521	1.8538	ug/L	99
32) cis-1,2-Dichloroethene	9.17	96	18238	1.8309	ug/L	91
33) Chloroform	9.37	83	33543	1.8156	ug/L	99
34) 1-Bromopropane	9.52	122	2513	1.6014	ug/L	88
35) Bromochloromethane	9.60	130	9189	1.7632	ug/L	67
36) Tetrahydrofuran	9.63	42	22074	10.8529	ug/L #	66
38) 1,1,1-Trichloroethane	9.90	97	32268	1.7324	ug/L	96
39) Cyclohexane	9.93	56	23097	1.1362	ug/L #	65
40) 1,1-Dichloropropene	10.10	75	24988	1.8639	ug/L	99
41) Tert-Amyl-Methyl ether	10.21	73	459535	23.9029	ug/L	87
42) Carbon Tetrachloride	10.24	117	28935	1.6512	ug/L	99

(#) = qualifier out of range (m) = manual integration
8M412223.D 8260WT.M Fri May 13 15:18:36 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412223.D Vial: 6
Acq On : 13 May 2016 14:56 Operator: TMB
Sample : WG568561-05 2ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:18:34 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit	Qvalue
45)	1,2-Dichloroethane	10.42	62	22142	1.4270 ug/L #	83
46)	Benzene	10.46	78	67028	1.9629 ug/L	83
47)	Trichloroethene	11.21	130	17287	1.7543 ug/L	96
48)	Methylcyclohexane	11.30	83	26050	1.8907 ug/L	84
49)	1,2-Dichloropropane	11.42	63	14582	1.4670 ug/L	77
50)	Bromodichloromethane	11.73	83	25794	1.8651 ug/L	93
51)	1,4-Dioxane	11.73	88	1282	32.3515 ug/L	74
52)	Dibromomethane	11.81	93	8575	1.8325 ug/L	97
53)	2-Chloroethyl Vinyl Ether	12.02	63	6234	1.4979 ug/L	100
54)	4-Methyl-2-Pentanone	12.06	58	2139	0.9416 ug/L	84
55)	cis-1,3-Dichloropropene	12.37	75	24153	1.7353 ug/L	98
56)	Dimethyl Disulfide	12.62	79	12288	1.6851 ug/L	99
59)	Toluene	12.79	91	72944	2.2423 ug/L	95
60)	Ethyl Methacrylate	12.88	69	12716	2.0832 ug/L	71
62)	trans-1,3-Dichloropropene	12.96	75	22163	1.9527 ug/L	100
63)	1,1,2-Trichloroethane	13.18	97	10537	2.0412 ug/L	99
64)	2-Hexanone	13.13	58	2200	1.1873 ug/L #	65
65)	1,3-Dichloropropane	13.49	76	20212	2.1112 ug/L	72
66)	Tetrachloroethene	13.60	164	14595	1.9811 ug/L	90
67)	Dibromochloromethane	13.87	129	14698	1.8061 ug/L	98
68)	1,2-Dibromoethane	14.12	107	10152	1.8869 ug/L	95
69)	1-Chlorohexane	14.21	91	21960	1.9818 ug/L	89
70)	Chlorobenzene	14.62	112	47376	2.0974 ug/L	81
71)	1,1,1,2-Tetrachloroethane	14.66	131	17526	1.9910 ug/L	97
72)	Ethylbenzene	14.66	106	25628	1.9996 ug/L	92
73)	m-,p-Xylene	14.74	106	60087	4.0693 ug/L	84
74)	o-Xylene	15.31	106	28289	1.9017 ug/L	79
75)	Styrene	15.35	104	47644	2.0086 ug/L	92
76)	Bromoform	15.84	173	9145	1.9585 ug/L	95
77)	Isopropylbenzene	15.73	105	81301	2.1557 ug/L	97
79)	1,1,2,2-Tetrachloroethane	15.95	83	11538	2.0194 ug/L	95
81)	1,2,3-Trichloropropane	16.13	110	3630	1.9770 ug/L	94
82)	trans-1,4-Dichloro-2-Buten	16.19	53	2699	0.9698 ug/L #	36
83)	n-Propylbenzene	16.24	91	97233	2.2026 ug/L	96
84)	Bromobenzene	16.37	156	19174	2.0834 ug/L	92
85)	1,3,5-Trimethylbenzene	16.42	105	67395	2.0631 ug/L	97
86)	2-Chlorotoluene	16.51	91	67110	2.3739 ug/L	94
87)	4-Chlorotoluene	16.56	91	59154	2.0313 ug/L	92
88)	a-Methylstyrene	16.84	118	29415	1.8275 ug/L	86
89)	tert-Butylbenzene	16.89	134	12716	1.9647 ug/L	67
90)	1,2,4-Trimethylbenzene	16.95	105	71669	2.1435 ug/L	94
91)	sec-Butylbenzene	17.17	105	81038	2.0632 ug/L	98
92)	p-Isopropyltoluene	17.32	119	70170	2.0924 ug/L	92
93)	1,3-Dichlorobenzene	17.52	146	37289	2.0413 ug/L	93
94)	1,4-Dichlorobenzene	17.64	146	36528	2.0034 ug/L #	73
95)	n-Butylbenzene	17.85	91	67823	2.1262 ug/L	96
96)	1,2-Dichlorobenzene	18.14	146	32322	2.0190 ug/L	92
97)	1,2-Dibromo-3-Chloropropan	19.13	75	1955	1.5624 ug/L	94
98)	1,2,4-Trichlorobenzene	20.28	180	23558	1.9000 ug/L	98
99)	Hexachlorobutadiene	20.43	225	11066	1.9286 ug/L	99
100)	Naphthalene	20.65	128	36066	1.9130 ug/L	99
101)	1,2,3-Trichlorobenzene	20.96	180	19765	1.8738 ug/L	95

(#) = qualifier out of range (m) = manual integration
8M412223.D 8260WT.M Fri May 13 15:18:37 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412223.D Vial: 6
Acq On : 13 May 2016 14:56 Operator: TMB
Sample : WG568561-05 2ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:18 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration



8M412223.D 8260WT.M Fri May 13 15:18:37 2016 Page 3

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412223.D Vial: 6
Acq On : 13 May 2016 14:56 Operator: TMB
Sample : WG568561-05 2ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	2.0000	1.8958	5.2	100	0.00
3	P	Chloromethane	2.0000	2.1033	-5.2	100	0.00
4	C	Vinyl Chloride	2.0000	2.0431	-2.2	100	0.00
5	T	1,3-Butadiene	2.0000	1.7872	10.6	100	0.02
6	T	Bromomethane	2.0000	2.0126	-0.6	100	0.00
7	T	Chloroethane	2.0000	1.9127	4.4	100	0.00
8	T	Trichlorofluoromethane	2.0000	2.0686	-3.4	100	0.00
9	T	Diethyl ether	25.0000	26.7567	-7.0	100	0.00
10	T	Isoprene	2.0000	2.0394	-2.0	100	0.00
11	T	Acrolein	12.5000	13.4606	-7.7	100	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	2.0000	2.1182	-5.9	100	0.02
13	T	Acetone	-1.0000	1.7255	0.0	100	-0.01
14	C	1,1-Dichloroethene	2.0000	2.0070	-0.4	100	0.00
15	T	Tert-Butyl Alcohol	50.0000	52.7840	-5.6	100	0.00
16	T	Dimethyl Sulfide	2.0000	2.0975	-4.9	100	0.00
17	T	Iodomethane	2.0000	3.0680	-53.4#	100	0.00
18	T	Methyl acetate	-1.0000	2.1281	0.0	100	0.00
19	T	Methylene Chloride	2.0000	2.0590	-3.0	100	0.00
20	T	Carbon Disulfide	2.0000	2.1789	-8.9	100	0.00
21	T	Acrylonitrile	12.5000	12.1237	3.0	100	0.00
22	T	Methyl Tert Butyl Ether	2.0000	1.9505	2.5	100	0.00
23	T	trans-1,2-Dichloroethene	2.0000	1.9603	2.0	100	0.00
24	T	n-Hexane	-1.0000	2.1181	0.0	100	0.00
25	T	Diisopropyl ether	25.0000	26.2305	-4.9	100	0.00
26	T	Vinyl Acetate	2.0000	1.9284	3.6	100	0.00
27	P	1,1-Dichloroethane	2.0000	2.1072	-5.4	100	0.00
28	T	Ethyl-Tert-Butyl ether	25.0000	26.0139	-4.1	100	0.00
29	T	2-Butanone	-1.0000	1.7396	0.0	100	0.02
30	T	Propionitrile	25.0000	24.8668	0.5	100	0.00
31	T	2,2-Dichloropropane	2.0000	2.1378	-6.9	100	0.00
32	T	cis-1,2-Dichloroethene	2.0000	2.0306	-1.5	100	0.00
33	C	Chloroform	2.0000	1.9679	1.6	100	0.00
34		1-Bromopropane	2.0000	1.9533	2.3	100	0.00
35	T	Bromochloromethane	2.0000	1.9951	0.2	100	0.00
36	T	Tetrahydrofuran	25.0000	25.9117	-3.6	100	0.00
37	S	Dibromofluoromethane	1.0000	0.9708	2.9	100	0.00
38	T	1,1,1-Trichloroethane	2.0000	1.9930	0.3	100	0.00
39	T	Cyclohexane	2.0000	2.0605	-3.0	100	0.00
40	T	1,1-Dichloropropene	2.0000	2.0248	-1.2	100	0.00
41	T	Tert-Amyl-Methyl ether	25.0000	26.0627	-4.3	100	0.00
42	T	Carbon Tetrachloride	2.0000	2.0126	-0.6	100	0.00
43	S	1,2-Dichloroethane-d4	1.0000	1.0740	-7.4	100	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	2.0000	2.0183	-0.9	100	0.00
46	T	Benzene	2.0000	2.0854	-4.3	100	0.00
47	T	Trichloroethene	2.0000	2.0166	-0.8	100	0.00
48	T	Methylcyclohexane	-1.0000	2.0003	0.0	100	0.00
49	C	1,2-Dichloropropane	2.0000	2.0380	-1.9	100	0.00
50	T	Bromodichloromethane	2.0000	2.0511	-2.6	100	0.00
51	T	1,4-Dioxane	-1.0000	35.6688	0.0	100	0.02
52	T	Dibromomethane	2.0000	2.0246	-1.2	100	0.00
53	T	2-Chloroethyl Vinyl Ether	2.0000	2.0271	-1.4	100	0.00
54	T	4-Methyl-2-Pentanone	-1.0000	1.7027	0.0	100	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412223.D Vial: 6
Acq On : 13 May 2016 14:56 Operator: TMB
Sample : WG568561-05 2ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	2.0000	1.8869	5.7	100	0.00
56	T	Dimethyl Disulfide	-1.0000	1.7986	0.0	100	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	1.0000	0.9896	1.0	100	0.00
59	C	Toluene	2.0000	2.1628	-8.1	100	0.00
60	T	Ethyl Methacrylate	2.0000	1.9633	1.8	100	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	2.0000	1.9609	2.0	100	0.00
63	T	1,1,2-Trichloroethane	2.0000	1.9800	1.0	100	0.00
64	T	2-Hexanone	-1.0000	1.9672	0.0	100	0.00
65	T	1,3-Dichloropropane	2.0000	2.1332	-6.7	100	0.00
66	T	Tetrachloroethene	2.0000	2.0198	-1.0	100	0.00
67	T	Dibromochloromethane	2.0000	1.8915	5.4	100	0.00
68	T	1,2-Dibromoethane	2.0000	1.9141	4.3	100	0.00
69	T	1-Chlorohexane	2.0000	1.9075	4.6	100	0.00
70	P	Chlorobenzene	2.0000	2.0831	-4.2	100	0.00
71	T	1,1,1,2-Tetrachloroethane	2.0000	1.9892	0.5	100	0.00
72	C	Ethylbenzene	2.0000	2.0596	-3.0	100	0.00
73	T	m-,p-Xylene	4.0000	4.1170	-2.9	100	0.00
74	T	o-Xylene	2.0000	1.9825	0.9	100	0.00
75	T	Styrene	2.0000	2.0358	-1.8	100	0.00
76	P	Bromoform	2.0000	1.9841	0.8	100	0.00
77	T	Isopropylbenzene	2.0000	2.1448	-7.2	100	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
79	P	1,1,2,2-Tetrachloroethane	2.0000	2.0066	-0.3	100	0.00
80	S	p-Bromofluorobenzene	1.0000	1.0184	-1.8	100	0.00
81	T	1,2,3-Trichloropropane	2.0000	2.0501	-2.5	100	0.00
82	T	trans-1,4-Dichloro-2-Butene	2.0000	2.0915	-4.6	100	0.00
83	T	n-Propylbenzene	2.0000	2.1757	-8.8	100	0.00
84	T	Bromobenzene	2.0000	2.0859	-4.3	100	0.00
85	T	1,3,5-Trimethylbenzene	2.0000	2.0703	-3.5	100	0.00
86	T	2-Chlorotoluene	2.0000	2.1509	-7.5	100	0.00
87	T	4-Chlorotoluene	2.0000	2.0989	-4.9	100	0.00
88	T	a-Methylstyrene	-1.0000	1.9045	0.0	100	0.00
89	T	tert-Butylbenzene	2.0000	1.9967	0.2	100	0.00
90	T	1,2,4-Trimethylbenzene	2.0000	2.1411	-7.1	100	0.00
91	T	sec-Butylbenzene	2.0000	2.1161	-5.8	100	0.00
92	T	p-Isopropyltoluene	2.0000	2.1579	-7.9	100	0.00
93	T	1,3-Dichlorobenzene	2.0000	2.1067	-5.3	100	0.00
94	T	1,4-Dichlorobenzene	2.0000	2.0470	-2.4	100	0.00
95	T	n-Butylbenzene	2.0000	2.1098	-5.5	100	0.00
96	T	1,2-Dichlorobenzene	2.0000	2.0532	-2.7	100	0.00
97	T	1,2-Dibromo-3-Chloropropane	2.0000	1.6938	15.3	100	0.00
98	T	1,2,4-Trichlorobenzene	2.0000	2.0021	-0.1	100	0.00
99	T	Hexachlorobutadiene	2.0000	2.0128	-0.6	100	0.00
100	T	Naphthalene	2.0000	2.0630	-3.2	100	0.00
101	T	1,2,3-Trichlorobenzene	2.0000	1.9910	0.4	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412224.D Vial: 7
Acq On : 13 May 2016 15:25 Operator: TMB
Sample : WG568561-06 5ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:47:41 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	768397	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	534372	25.00 ug/L	0.01
78) 1,4-Dichlorobenzene-d4	17.60	152	274234	25.00 ug/L	0.01

System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	19775	2.0496 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery =	8.20%#	
43) 1,2-Dichloroethane-d4	10.30	65	21200	1.9207 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery =	7.68%#	
58) Toluene-d8	12.68	98	65647	2.3758 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery =	9.52%#	
80) p-Bromofluorobenzene	16.08	95	26759	2.3838 ug/L	0.01
Spiked Amount 25.000	Range 86 - 115		Recovery =	9.52%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	55961	3.4750 ug/L	99
3) Chloromethane	3.62	50	34194	2.2538 ug/L	100
4) Vinyl Chloride	3.84	62	42506	4.1488 ug/L	99
5) 1,3-Butadiene	3.89	54	32006	3.1976 ug/L	93
6) Bromomethane	4.72	94	19582	3.4204 ug/L	94
7) Chloroethane	4.90	64	22000	3.2814 ug/L	84
8) Trichlorofluoromethane	5.37	101	77715	3.8212 ug/L	96
9) Diethyl ether	5.90	59	188516	30.2955 ug/L	79
10) Isoprene	5.94	67	44975	4.3218 ug/L	78
11) Acrolein	6.15	56	7712	13.9568 ug/L	92
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	36867	3.9591 ug/L	99
13) Acetone	6.24	43	3578	1.8967 ug/L	85
14) 1,1-Dichloroethene	6.47	61	59405	3.2870 ug/L	95
15) Tert-Butyl Alcohol	6.58	59	27833	64.0897 ug/L #	82
16) Dimethyl Sulfide	6.73	62	21102	3.1597 ug/L	77
17) Iodomethane	6.99	142	3667	1.8964 ug/L	95
18) Methyl acetate	7.00	43	11500	2.0398 ug/L #	86
19) Methylene Chloride	7.24	84	37533	4.3823 ug/L	52
20) Carbon Disulfide	7.29	76	123060	4.7097 ug/L	100
21) Acrylonitrile	7.43	53	25653	10.5804 ug/L	99
22) Methyl Tert Butyl Ether	7.47	73	80005	4.2289 ug/L	86
23) trans-1,2-Dichloroethene	7.70	61	54124	3.3108 ug/L	94
24) n-Hexane	7.78	57	40758	2.7480 ug/L	82
25) Diisopropyl ether	8.12	45	841400	20.0337 ug/L #	91
26) Vinyl Acetate	8.30	43	37284	2.2113 ug/L #	90
27) 1,1-Dichloroethane	8.32	63	66711	3.4116 ug/L	97
28) Ethyl-Tert-Butyl ether	8.70	59	907423	30.7852 ug/L	89
29) 2-Butanone	8.89	43	5359	1.7738 ug/L #	82
30) Propionitrile	8.99	54	17933	23.4726 ug/L	100
31) 2,2-Dichloropropane	9.10	77	73150	4.2462 ug/L	99
32) cis-1,2-Dichloroethene	9.17	96	41036	4.1950 ug/L	89
33) Chloroform	9.37	83	78227	4.3116 ug/L	99
34) 1-Bromopropane	9.52	122	6162	3.9986 ug/L	99
35) Bromochloromethane	9.60	130	20988	4.1009 ug/L #	62
36) Tetrahydrofuran	9.62	42	38098	19.0742 ug/L #	62
38) 1,1,1-Trichloroethane	9.91	97	76152	4.1633 ug/L	97
39) Cyclohexane	9.94	56	51439	2.5767 ug/L #	66
40) 1,1-Dichloropropene	10.11	75	56645	4.3027 ug/L	93
41) Tert-Amyl-Methyl ether	10.21	73	821180	43.4958 ug/L	87
42) Carbon Tetrachloride	10.25	117	67831	3.9416 ug/L	96

(#) = qualifier out of range (m) = manual integration
8M412224.D 8260WT.M Fri May 13 15:47:43 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412224.D Vial: 7
Acq On : 13 May 2016 15:25 Operator: TMB
Sample : WG568561-06 5ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:47:41 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
45)	1,2-Dichloroethane	10.43	62	51708	3.3934 ug/L		85
46)	Benzene	10.46	78	150204	4.4792 ug/L		86
47)	Trichloroethene	11.21	130	38500	3.9784 ug/L		94
48)	Methylcyclohexane	11.30	83	60897	4.5008 ug/L		81
49)	1,2-Dichloropropane	11.43	63	31773	3.2549 ug/L		77
50)	Bromodichloromethane	11.73	83	59127	4.3535 ug/L		94
51)	1,4-Dioxane	11.71	88	2719	69.8702 ug/L	#	62
52)	Dibromomethane	11.81	93	19711	4.2894 ug/L		96
53)	2-Chloroethyl Vinyl Ether	12.03	63	12946	3.1676 ug/L		98
54)	4-Methyl-2-Pentanone	12.05	58	5129	2.2991 ug/L	#	78
55)	cis-1,3-Dichloropropene	12.36	75	61169	4.4752 ug/L		94
56)	Dimethyl Disulfide	12.62	79	27966	3.9053 ug/L		98
59)	Toluene	12.78	91	161007	5.0316 ug/L		97
60)	Ethyl Methacrylate	12.88	69	28595	4.7624 ug/L		71
62)	trans-1,3-Dichloropropene	12.96	75	49791	4.4599 ug/L		97
63)	1,1,2-Trichloroethane	13.18	97	23678	4.6631 ug/L		99
64)	2-Hexanone	13.11	58	4499	2.4685 ug/L	#	97
65)	1,3-Dichloropropane	13.49	76	44838	4.7614 ug/L		73
66)	Tetrachloroethene	13.60	164	32715	4.5144 ug/L		97
67)	Dibromochloromethane	13.87	129	35009	4.3734 ug/L		98
68)	1,2-Dibromoethane	14.13	107	24247	4.5815 ug/L		99
69)	1-Chlorohexane	14.21	91	53599	4.9176 ug/L		84
70)	Chlorobenzene	14.62	112	107618	4.8437 ug/L		95
71)	1,1,1,2-Tetrachloroethane	14.65	131	38120	4.4025 ug/L		96
72)	Ethylbenzene	14.65	106	56905	4.5138 ug/L		91
73)	m-,p-Xylene	14.75	106	137390	9.4592 ug/L		86
74)	o-Xylene	15.31	106	65650	4.4866 ug/L		85
75)	Styrene	15.35	104	111433	4.7758 ug/L		91
76)	Bromoform	15.84	173	19346	4.2121 ug/L		97
77)	Isopropylbenzene	15.73	105	182523	4.9200 ug/L		96
79)	1,1,2,2-Tetrachloroethane	15.95	83	26655	4.8730 ug/L		99
81)	1,2,3-Trichloropropane	16.13	110	8840	5.0289 ug/L		81
82)	trans-1,4-Dichloro-2-Buten	16.19	53	6584	2.4712 ug/L	#	51
83)	n-Propylbenzene	16.24	91	222196	5.2576 ug/L		96
84)	Bromobenzene	16.37	156	42665	4.8423 ug/L		95
85)	1,3,5-Trimethylbenzene	16.43	105	155714	4.9790 ug/L		93
86)	2-Chlorotoluene	16.52	91	148764	5.4968 ug/L		94
87)	4-Chlorotoluene	16.56	91	135679	4.8666 ug/L		94
88)	a-Methylstyrene	16.84	118	68945	4.4743 ug/L		91
89)	tert-Butylbenzene	16.89	134	29069	4.6914 ug/L		77
90)	1,2,4-Trimethylbenzene	16.95	105	161016	5.0301 ug/L		94
91)	sec-Butylbenzene	17.17	105	191273	5.0866 ug/L		94
92)	p-Isopropyltoluene	17.32	119	159232	4.9597 ug/L		93
93)	1,3-Dichlorobenzene	17.52	146	83896	4.7972 ug/L		95
94)	1,4-Dichlorobenzene	17.64	146	82462	4.7242 ug/L		86
95)	n-Butylbenzene	17.85	91	158787	5.1996 ug/L		96
96)	1,2-Dichlorobenzene	18.14	146	72524	4.7321 ug/L		92
97)	1,2-Dibromo-3-Chloropropan	19.13	75	5629	4.6990 ug/L		76
98)	1,2,4-Trichlorobenzene	20.28	180	53855	4.5369 ug/L		98
99)	Hexachlorobutadiene	20.43	225	24504	4.4608 ug/L		93
100)	Naphthalene	20.65	128	79806	4.4217 ug/L		98
101)	1,2,3-Trichlorobenzene	20.96	180	44150	4.3720 ug/L		99

(#) = qualifier out of range (m) = manual integration
8M412224.D 8260WT.M Fri May 13 15:47:44 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412224.D Vial: 7
Acq On : 13 May 2016 15:25 Operator: TMB
Sample : WG568561-06 5ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:47 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412224.D Vial: 7
Acq On : 13 May 2016 15:25 Operator: TMB
Sample : WG568561-06 5ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	5.0000	4.3438	13.1	100	0.00
3	P	Chloromethane	5.0000	4.5471	9.1	100	0.00
4	C	Vinyl Chloride	5.0000	4.7641	4.7	100	0.00
5	T	1,3-Butadiene	5.0000	4.2777	14.4	100	0.00
6	T	Bromomethane	5.0000	4.1880	16.2	100	-0.01
7	T	Chloroethane	5.0000	4.4813	10.4	100	0.00
8	T	Trichlorofluoromethane	5.0000	4.6812	6.4	100	0.00
9	T	Diethyl ether	50.0000	47.5652	4.9	100	0.00
10	T	Isoprene	5.0000	4.4611	10.8	100	0.00
11	T	Acrolein	25.0000	21.7328	13.1	100	0.02
12	T	1,1,2-Trichloro-1,2,2-Trifl	5.0000	4.7244	5.5	100	0.00
13	T	Acetone	5.0000	4.1387	17.2	100	0.00
14	C	1,1-Dichloroethene	5.0000	4.7810	4.4	100	0.00
15	T	Tert-Butyl Alcohol	100.0000	100.4650	-0.5	100	0.00
16	T	Dimethyl Sulfide	5.0000	4.7499	5.0	100	0.00
17	T	Iodomethane	5.0000	3.7634	24.7	100	0.00
18	T	Methyl acetate	5.0000	4.9211	1.6	100	0.00
19	T	Methylene Chloride	5.0000	4.7467	5.1	100	0.00
20	T	Carbon Disulfide	5.0000	4.7328	5.3	100	0.00
21	T	Acrylonitrile	25.0000	22.8698	8.5	100	0.00
22	T	Methyl Tert Butyl Ether	5.0000	4.6750	6.5	100	0.00
23	T	trans-1,2-Dichloroethene	5.0000	4.7415	5.2	100	0.00
24	T	n-Hexane	5.0000	4.8084	3.8	100	0.00
25	T	Diisopropyl ether	50.0000	47.2872	5.4	100	0.00
26	T	Vinyl Acetate	5.0000	4.5766	8.5	100	0.00
27	P	1,1-Dichloroethane	5.0000	4.7058	5.9	100	0.00
28	T	Ethyl-Tert-Butyl ether	50.0000	47.5984	4.8	100	0.00
29	T	2-Butanone	5.0000	4.2191	15.6	100	0.00
30	T	Propionitrile	50.0000	48.4475	3.1	100	0.00
31	T	2,2-Dichloropropane	5.0000	4.8965	2.1	100	0.00
32	T	cis-1,2-Dichloroethene	5.0000	4.6525	7.0	100	0.00
33	C	Chloroform	5.0000	4.6735	6.5	100	0.00
34		1-Bromopropane	5.0000	4.8774	2.5	100	0.00
35	T	Bromochloromethane	5.0000	4.6403	7.2	100	0.00
36	T	Tetrahydrofuran	50.0000	45.5400	8.9	100	-0.01
37	S	Dibromofluoromethane	2.5000	2.3953	4.2	100	0.00
38	T	1,1,1-Trichloroethane	5.0000	4.7896	4.2	100	0.00
39	T	Cyclohexane	5.0000	4.6730	6.5	100	0.00
40	T	1,1-Dichloropropene	5.0000	4.6741	6.5	100	0.00
41	T	Tert-Amyl-Methyl ether	50.0000	47.6211	4.8	100	0.00
42	T	Carbon Tetrachloride	5.0000	4.8044	3.9	100	0.00
43	S	1,2-Dichloroethane-d4	2.5000	2.3949	4.2	100	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	5.0000	4.7996	4.0	100	0.00
46	T	Benzene	5.0000	4.7588	4.8	100	0.00
47	T	Trichloroethene	5.0000	4.5733	8.5	100	0.00
48	T	Methylcyclohexane	5.0000	4.7616	4.8	100	0.00
49	C	1,2-Dichloropropane	5.0000	4.5219	9.6	100	0.00
50	T	Bromodichloromethane	5.0000	4.7878	4.2	100	0.00
51	T	1,4-Dioxane	100.0000	77.0346	23.0	100	0.00
52	T	Dibromomethane	5.0000	4.7390	5.2	100	0.00
53	T	2-Chloroethyl Vinyl Ether	5.0000	4.2867	14.3	100	0.00
54	T	4-Methyl-2-Pentanone	5.0000	4.1576	16.8	100	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412224.D Vial: 7
Acq On : 13 May 2016 15:25 Operator: TMB
Sample : WG568561-06 5ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	5.0000	4.8662	2.7	100	0.00
56	T	Dimethyl Disulfide	5.0000	4.1683	16.6	100	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	2.5000	2.3473	6.1	100	0.00
59	C	Toluene	5.0000	4.8533	2.9	100	0.00
60	T	Ethyl Methacrylate	5.0000	4.4883	10.2	100	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	5.0000	4.4786	10.4	100	0.00
63	T	1,1,2-Trichloroethane	5.0000	4.5231	9.5	100	0.00
64	T	2-Hexanone	5.0000	4.0898	18.2	100	0.00
65	T	1,3-Dichloropropane	5.0000	4.8109	3.8	100	0.00
66	T	Tetrachloroethene	5.0000	4.6028	7.9	100	0.00
67	T	Dibromochloromethane	5.0000	4.5802	8.4	100	0.00
68	T	1,2-Dibromoethane	5.0000	4.6475	7.1	100	0.00
69	T	1-Chlorohexane	5.0000	4.7331	5.3	100	0.00
70	P	Chlorobenzene	5.0000	4.8107	3.8	100	0.00
71	T	1,1,1,2-Tetrachloroethane	5.0000	4.3984	12.0	100	0.00
72	C	Ethylbenzene	5.0000	4.6492	7.0	100	0.00
73	T	m-,p-Xylene	10.0000	9.5700	4.3	100	0.00
74	T	o-Xylene	5.0000	4.6773	6.5	100	0.00
75	T	Styrene	5.0000	4.8405	3.2	100	0.00
76	P	Bromoform	5.0000	4.2672	14.7	100	0.00
77	T	Isopropylbenzene	5.0000	4.8951	2.1	100	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
79	P	1,1,2,2-Tetrachloroethane	5.0000	4.8422	3.2	100	0.00
80	S	p-Bromofluorobenzene	2.5000	2.3565	5.7	100	0.00
81	T	1,2,3-Trichloropropane	5.0000	5.2150	-4.3	100	0.00
82	T	trans-1,4-Dichloro-2-Butene	5.0000	4.3781	12.4	100	0.00
83	T	n-Propylbenzene	5.0000	5.1935	-3.9	100	0.00
84	T	Bromobenzene	5.0000	4.8483	3.0	100	0.00
85	T	1,3,5-Trimethylbenzene	5.0000	4.9964	0.1	100	0.00
86	T	2-Chlorotoluene	5.0000	4.9803	0.4	100	0.00
87	T	4-Chlorotoluene	5.0000	5.0286	-0.6	100	0.00
88	T	a-Methylstyrene	5.0000	4.6629	6.7	100	0.00
89	T	tert-Butylbenzene	5.0000	4.7677	4.6	100	0.00
90	T	1,2,4-Trimethylbenzene	5.0000	5.0245	-0.5	100	0.00
91	T	sec-Butylbenzene	5.0000	5.2170	-4.3	100	0.00
92	T	p-Isopropyltoluene	5.0000	5.1150	-2.3	100	0.00
93	T	1,3-Dichlorobenzene	5.0000	4.9511	1.0	100	0.00
94	T	1,4-Dichlorobenzene	5.0000	4.8270	3.5	100	0.00
95	T	n-Butylbenzene	5.0000	5.1595	-3.2	100	0.00
96	T	1,2-Dichlorobenzene	5.0000	4.8122	3.8	100	0.00
97	T	1,2-Dibromo-3-Chloropropane	5.0000	5.0942	-1.9	100	0.00
98	T	1,2,4-Trichlorobenzene	5.0000	4.7808	4.4	100	0.00
99	T	Hexachlorobutadiene	5.0000	4.6555	6.9	100	0.00
100	T	Naphthalene	5.0000	4.7684	4.6	100	0.00
101	T	1,2,3-Trichlorobenzene	5.0000	4.6455	7.1	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412224.D 8260WT.M Sat May 14 18:08:37 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412225.D Vial: 8
Acq On : 13 May 2016 15:54 Operator: TMB
Sample : WG568561-07 20ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:16:32 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	770027	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	531490	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	283594	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	86295	10.9569 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 43.84%#		
43) 1,2-Dichloroethane-d4	10.30	65	96239	11.3450 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 45.36%#		
58) Toluene-d8	12.68	98	294491	10.9223 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 43.68%#		
80) p-Bromofluorobenzene	16.08	95	121459	10.5490 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 42.20%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	288235	23.7135 ug/L	99
3) Chloromethane	3.62	50	155712	20.3784 ug/L	98
4) Vinyl Chloride	3.84	62	184250	20.9368 ug/L	95
5) 1,3-Butadiene	3.89	54	149348	22.7750 ug/L	96
6) Bromomethane	4.73	94	100098	25.0143 ug/L	99
7) Chloroethane	4.89	64	100876	20.9404 ug/L	91
8) Trichlorofluoromethane	5.37	101	351299	21.2512 ug/L	98
9) Diethyl ether	5.90	59	324317	80.8265 ug/L	78
10) Isoprene	5.94	67	205842	20.3622 ug/L	77
11) Acrolein	6.14	56	12049	34.8201 ug/L	90
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	159996	20.3859 ug/L	97
13) Acetone	6.25	43	18459	25.7405 ug/L	98
14) 1,1-Dichloroethene	6.47	61	256241	20.5065 ug/L	92
15) Tert-Butyl Alcohol	6.57	59	43922	153.5712 ug/L	93
16) Dimethyl Sulfide	6.73	62	89456	20.6087 ug/L	83
17) Iodomethane	6.99	142	40622	61.0474 ug/L	93
18) Methyl acetate	7.00	43	46598	20.2171 ug/L #	85
19) Methylene Chloride	7.25	84	161897	20.1338 ug/L	52
20) Carbon Disulfide	7.30	76	538090	20.0802 ug/L	98
21) Acrylonitrile	7.43	53	46305	46.9889 ug/L	100
22) Methyl Tert Butyl Ether	7.47	73	361116	22.0137 ug/L	86
23) trans-1,2-Dichloroethene	7.70	61	234754	20.3175 ug/L	90
24) n-Hexane	7.78	57	172498	21.1165 ug/L	84
25) Diisopropyl ether	8.12	45	1447822	80.3755 ug/L #	91
26) Vinyl Acetate	8.30	43	166501	21.7017 ug/L #	88
27) 1,1-Dichloroethane	8.33	63	300321	21.3849 ug/L	97
28) Ethyl-Tert-Butyl ether	8.70	59	1579265	83.0733 ug/L	89
29) 2-Butanone	8.88	43	25681	23.9099 ug/L	87
30) Propionitrile	8.98	54	30953	89.5673 ug/L	94
31) 2,2-Dichloropropane	9.10	77	309575	20.8527 ug/L	100
32) cis-1,2-Dichloroethene	9.17	96	181973	20.4488 ug/L	89
33) Chloroform	9.37	83	348925	20.5535 ug/L	96
34) 1-Bromopropane	9.52	122	26723	22.7206 ug/L	96
35) Bromochloromethane	9.61	130	94952	21.3181 ug/L	65
36) Tetrahydrofuran	9.63	42	63066	70.5772 ug/L #	57
38) 1,1,1-Trichloroethane	9.91	97	335486	21.3640 ug/L	99
39) Cyclohexane	9.94	56	221040	19.9984 ug/L #	64
40) 1,1-Dichloropropene	10.11	75	248373	20.4717 ug/L	96
41) Tert-Amyl-Methyl ether	10.21	73	1418183	81.7883 ug/L	88
42) Carbon Tetrachloride	10.25	117	296471	21.3449 ug/L	98

(#) = qualifier out of range (m) = manual integration
8M412225.D 8260WT.M Fri May 13 16:16:35 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412225.D Vial: 8
Acq On : 13 May 2016 15:54 Operator: TMB
Sample : WG568561-07 20ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:16:32 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc	Unit		Qvalue
45)	1,2-Dichloroethane	10.43	62	228665	21.6403	ug/L		85
46)	Benzene	10.46	78	668449	20.3769	ug/L		86
47)	Trichloroethene	11.21	130	174107	20.7980	ug/L		96
48)	Methylcyclohexane	11.29	83	262275	21.4887	ug/L		82
49)	1,2-Dichloropropane	11.43	63	146868	21.1883	ug/L	#	72
50)	Bromodichloromethane	11.73	83	262192	21.9986	ug/L		98
51)	1,4-Dioxane	11.71	88	5351	196.3837	ug/L		76
52)	Dibromomethane	11.81	93	86951	21.0734	ug/L		98
53)	2-Chloroethyl Vinyl Ether	12.03	63	63586	22.4600	ug/L		91
54)	4-Methyl-2-Pentanone	12.05	58	25754	25.0531	ug/L	#	73
55)	cis-1,3-Dichloropropene	12.36	75	265439	21.9841	ug/L		99
56)	Dimethyl Disulfide	12.63	79	134192	23.9412	ug/L		100
59)	Toluene	12.78	91	700473	20.1722	ug/L		99
60)	Ethyl Methacrylate	12.88	69	133361	23.2046	ug/L		71
62)	trans-1,3-Dichloropropene	12.96	75	238585	23.2548	ug/L		99
63)	1,1,2-Trichloroethane	13.18	97	107901	21.0216	ug/L		99
64)	2-Hexanone	13.11	58	22561	25.2093	ug/L	#	85
65)	1,3-Dichloropropane	13.49	76	198804	22.2491	ug/L		74
66)	Tetrachloroethene	13.60	164	141791	20.0055	ug/L		94
67)	Dibromochloromethane	13.87	129	158465	22.0695	ug/L		99
68)	1,2-Dibromoethane	14.13	107	110085	22.3366	ug/L		100
69)	1-Chlorohexane	14.21	91	229753	20.3591	ug/L		83
70)	Chlorobenzene	14.62	112	467492	20.4613	ug/L		97
71)	1,1,1,2-Tetrachloroethane	14.65	131	183079	22.2095	ug/L		96
72)	Ethylbenzene	14.65	106	249016	20.2289	ug/L		90
73)	m-,p-Xylene	14.75	106	606381	41.1308	ug/L		88
74)	o-Xylene	15.30	106	294106	21.2106	ug/L		90
75)	Styrene	15.35	104	492781	21.8207	ug/L		89
76)	Bromoform	15.83	173	95911	23.8070	ug/L		99
77)	Isopropylbenzene	15.73	105	806678	20.7280	ug/L		97
79)	1,1,2,2-Tetrachloroethane	15.95	83	123253	22.7779	ug/L		95
81)	1,2,3-Trichloropropane	16.14	110	36712	22.1354	ug/L		94
82)	trans-1,4-Dichloro-2-Buten	16.18	53	33702	27.6513	ug/L	#	56
83)	n-Propylbenzene	16.24	91	962196	20.3511	ug/L		97
84)	Bromobenzene	16.37	156	190566	21.2033	ug/L		94
85)	1,3,5-Trimethylbenzene	16.43	105	692633	20.9615	ug/L		94
86)	2-Chlorotoluene	16.51	91	649077	19.9155	ug/L		96
87)	4-Chlorotoluene	16.56	91	595446	20.3225	ug/L		95
88)	a-Methylstyrene	16.83	118	318960	22.3680	ug/L		92
89)	tert-Butylbenzene	16.90	134	129440	20.4705	ug/L		83
90)	1,2,4-Trimethylbenzene	16.95	105	707749	20.4783	ug/L		94
91)	sec-Butylbenzene	17.17	105	818359	20.1802	ug/L		97
92)	p-Isopropyltoluene	17.32	119	677859	19.7492	ug/L		96
93)	1,3-Dichlorobenzene	17.52	146	364065	20.5126	ug/L		92
94)	1,4-Dichlorobenzene	17.64	146	363066	19.9874	ug/L		92
95)	n-Butylbenzene	17.85	91	681055	20.0187	ug/L		97
96)	1,2-Dichlorobenzene	18.14	146	318612	19.9358	ug/L		91
97)	1,2-Dibromo-3-Chloropropan	19.13	75	23070	21.6419	ug/L		93
98)	1,2,4-Trichlorobenzene	20.28	180	242455	20.8863	ug/L		98
99)	Hexachlorobutadiene	20.43	225	110384	20.2523	ug/L		96
100)	Naphthalene	20.65	128	369849	21.4526	ug/L		99
101)	1,2,3-Trichlorobenzene	20.96	180	198876	19.8356	ug/L		98

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412225.D Vial: 8
Acq On : 13 May 2016 15:54 Operator: TMB
Sample : WG568561-07 20ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:16 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412226.D Vial: 9
Acq On : 13 May 2016 16:23 Operator: TMB
Sample : WG568561-08 50ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:45:30 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	778260	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	539116	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	288225	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.67	111	211053	26.5139 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 106.04%		
43) 1,2-Dichloroethane-d4	10.31	65	229348	26.7504 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 107.00%		
58) Toluene-d8	12.68	98	718491	26.2709 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 105.08%		
80) p-Bromofluorobenzene	16.07	95	292755	25.0180 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 100.08%		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.17	85	682745	55.5761 ug/L	100
3) Chloromethane	3.62	50	372858	48.2806 ug/L	98
4) Vinyl Chloride	3.85	62	453905	51.0328 ug/L	96
5) 1,3-Butadiene	3.88	54	390501	58.9199 ug/L	99
6) Bromomethane	4.74	94	260493	64.4080 ug/L	97
7) Chloroethane	4.89	64	244336	50.1840 ug/L	91
8) Trichlorofluoromethane	5.38	101	837676	50.1376 ug/L	99
9) Diethyl ether	5.90	59	389327	96.0019 ug/L	79
10) Isoprene	5.94	67	498060	48.7477 ug/L	76
11) Acrolein	6.13	56	17644	50.4496 ug/L	99
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	387643	48.8691 ug/L	97
13) Acetone	6.25	43	45901	63.3305 ug/L	97
14) 1,1-Dichloroethene	6.46	61	619744	49.0724 ug/L	94
15) Tert-Butyl Alcohol	6.58	59	53656	185.6211 ug/L	91
16) Dimethyl Sulfide	6.73	62	221107	50.3993 ug/L	82
17) Iodomethane	6.98	142	157872	234.7428 ug/L	94
18) Methyl acetate	7.00	43	114759	49.2629 ug/L #	81
19) Methylene Chloride	7.25	84	389653	47.9453 ug/L	52
20) Carbon Disulfide	7.29	76	1269941	46.8897 ug/L	99
21) Acrylonitrile	7.43	53	59882	60.1236 ug/L	97
22) Methyl Tert Butyl Ether	7.46	73	901668	54.3843 ug/L	86
23) trans-1,2-Dichloroethene	7.70	61	571507	48.9396 ug/L	93
24) n-Hexane	7.79	57	419656	50.8290 ug/L	83
25) Diisopropyl ether	8.13	45	1781012	97.8265 ug/L #	90
26) Vinyl Acetate	8.29	43	428708	55.2866 ug/L #	89
27) 1,1-Dichloroethane	8.32	63	722574	50.9078 ug/L	96
28) Ethyl-Tert-Butyl ether	8.71	59	1955100	101.7552 ug/L	89
29) 2-Butanone	8.88	43	66579	61.3316 ug/L #	82
30) Propionitrile	8.98	54	38235	109.4684 ug/L	95
31) 2,2-Dichloropropane	9.10	77	768496	51.2175 ug/L	99
32) cis-1,2-Dichloroethene	9.17	96	439590	48.8754 ug/L	91
33) Chloroform	9.38	83	844564	49.2230 ug/L	97
34) 1-Bromopropane	9.51	122	63442	53.3693 ug/L	99
35) Bromochloromethane	9.60	130	230119	51.1185 ug/L	66
36) Tetrahydrofuran	9.64	42	79437	87.9575 ug/L #	62
38) 1,1,1-Trichloroethane	9.90	97	817606	51.5150 ug/L	99
39) Cyclohexane	9.93	56	535882	47.9707 ug/L #	65
40) 1,1-Dichloropropene	10.10	75	615439	50.1899 ug/L	96
41) Tert-Amyl-Methyl ether	10.21	73	1769381	100.9628 ug/L	88
42) Carbon Tetrachloride	10.24	117	724058	51.5782 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M412226.D 8260WT.M Fri May 13 16:45:32 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412226.D Vial: 9
Acq On : 13 May 2016 16:23 Operator: TMB
Sample : WG568561-08 50ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:45:30 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
45)	1,2-Dichloroethane	10.42	62	561107	52.5401 ug/L		85
46)	Benzene	10.46	78	1607407	48.4816 ug/L		86
47)	Trichloroethene	11.22	130	424998	50.2312 ug/L		96
48)	Methylcyclohexane	11.30	83	635339	51.5040 ug/L		83
49)	1,2-Dichloropropane	11.42	63	351908	50.2320 ug/L	#	71
50)	Bromodichloromethane	11.72	83	654943	54.3701 ug/L		97
51)	1,4-Dioxane	11.71	88	7573	274.9918 ug/L		73
52)	Dibromomethane	11.82	93	209167	50.1574 ug/L		97
53)	2-Chloroethyl Vinyl Ether	12.02	63	162319	56.7281 ug/L		93
54)	4-Methyl-2-Pentanone	12.05	58	64928	62.4928 ug/L		80
55)	cis-1,3-Dichloropropene	12.36	75	666861	54.6463 ug/L		98
56)	Dimethyl Disulfide	12.62	79	343446	60.6260 ug/L		96
59)	Toluene	12.79	91	1676069	47.5846 ug/L		100
60)	Ethyl Methacrylate	12.88	69	338996	58.1503 ug/L		69
62)	trans-1,3-Dichloropropene	12.96	75	594749	57.1499 ug/L		100
63)	1,1,2-Trichloroethane	13.18	97	268432	51.5570 ug/L		98
64)	2-Hexanone	13.12	58	56859	62.6347 ug/L	#	83
65)	1,3-Dichloropropane	13.48	76	489226	53.9770 ug/L		75
66)	Tetrachloroethene	13.60	164	350332	48.7296 ug/L		95
67)	Dibromochloromethane	13.86	129	401149	55.0780 ug/L		100
68)	1,2-Dibromoethane	14.12	107	273580	54.7252 ug/L		99
69)	1-Chlorohexane	14.21	91	560883	48.9984 ug/L		82
70)	Chlorobenzene	14.63	112	1117812	48.2326 ug/L		96
71)	1,1,1,2-Tetrachloroethane	14.66	131	446892	53.4462 ug/L		99
72)	Ethylbenzene	14.66	106	603074	48.2980 ug/L		91
73)	m-,p-Xylene	14.74	106	1443780	96.5463 ug/L		90
74)	o-Xylene	15.31	106	709244	50.4265 ug/L		90
75)	Styrene	15.34	104	1212099	52.9133 ug/L		91
76)	Bromoform	15.84	173	238354	58.3272 ug/L		100
77)	Isopropylbenzene	15.73	105	1919993	48.6372 ug/L		97
79)	1,1,2,2-Tetrachloroethane	15.95	83	294266	53.5083 ug/L		97
81)	1,2,3-Trichloropropane	16.14	110	89279	52.9656 ug/L		93
82)	trans-1,4-Dichloro-2-Buten	16.19	53	86829	70.0956 ug/L	#	53
83)	n-Propylbenzene	16.24	91	2278994	47.4277 ug/L		98
84)	Bromobenzene	16.37	156	459959	50.3549 ug/L		93
85)	1,3,5-Trimethylbenzene	16.43	105	1661486	49.4745 ug/L		95
86)	2-Chlorotoluene	16.51	91	1611031	48.6367 ug/L		97
87)	4-Chlorotoluene	16.56	91	1341051	45.0344 ug/L		95
88)	a-Methylstyrene	16.83	118	772611	53.3111 ug/L		90
89)	tert-Butylbenzene	16.89	134	307503	47.8493 ug/L		81
90)	1,2,4-Trimethylbenzene	16.94	105	1679809	47.8234 ug/L		95
91)	sec-Butylbenzene	17.16	105	1925411	46.7166 ug/L		97
92)	p-Isopropyltoluene	17.32	119	1621383	46.4795 ug/L		95
93)	1,3-Dichlorobenzene	17.51	146	882656	48.9327 ug/L		93
94)	1,4-Dichlorobenzene	17.64	146	871210	47.1909 ug/L		92
95)	n-Butylbenzene	17.85	91	1608832	46.5296 ug/L		98
96)	1,2-Dichlorobenzene	18.14	146	768548	47.3159 ug/L		91
97)	1,2-Dibromo-3-Chloropropan	19.13	75	59586	54.9992 ug/L		86
98)	1,2,4-Trichlorobenzene	20.28	180	583880	49.4902 ug/L		100
99)	Hexachlorobutadiene	20.44	225	264699	47.7844 ug/L		98
100)	Naphthalene	20.65	128	894629	51.0579 ug/L		99
101)	1,2,3-Trichlorobenzene	20.96	180	475246	46.6387 ug/L		99

(#) = qualifier out of range (m) = manual integration
8M412226.D 8260WT.M Fri May 13 16:45:33 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412226.D Vial: 9
Acq On : 13 May 2016 16:23 Operator: TMB
Sample : WG568561-08 50ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:45 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412227.D Vial: 10
Acq On : 13 May 2016 16:52 Operator: TMB
Sample : WG568561-09 100ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:14:31 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:12:43 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	809533	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	559075	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	296193	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	460283	53.9056 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 215.64%#		
43) 1,2-Dichloroethane-d4	10.30	65	488116	52.5821 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 210.32%#		
58) Toluene-d8	12.68	98	1540572	52.8075 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 211.24%#		
80) p-Bromofluorobenzene	16.08	95	623735	51.1573 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 204.64%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	1463283	109.1068 ug/L	100
3) Chloromethane	3.62	50	778702	97.2380 ug/L	98
4) Vinyl Chloride	3.84	62	968312	103.4979 ug/L	95
5) 1,3-Butadiene	3.88	54	896569	120.4972 ug/L	95
6) Bromomethane	4.73	94	605228	129.8677 ug/L	99
7) Chloroethane	4.89	64	528153	103.4126 ug/L	90
8) Trichlorofluoromethane	5.37	101	1782726	101.4754 ug/L	99
9) Diethyl ether	5.90	59	828091	197.4770 ug/L	79
10) Isoprene	5.94	67	1080684	101.8037 ug/L	78
11) Acrolein	6.14	56	38655	109.5572 ug/L	98
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	833546	101.0906 ug/L	97
13) Acetone	6.24	43	97515	109.1944 ug/L	100
14) 1,1-Dichloroethene	6.46	61	1323999	100.6729 ug/L	93
15) Tert-Butyl Alcohol	6.57	59	114220	387.5328 ug/L	91
16) Dimethyl Sulfide	6.73	62	489571	106.5996 ug/L	82
17) Iodomethane	6.99	142	416890	244.5557 ug/L	97
18) Methyl acetate	7.00	43	252444	104.3161 ug/L #	81
19) Methylene Chloride	7.24	84	827241	98.5346 ug/L	53
20) Carbon Disulfide	7.29	76	2766110	99.3432 ug/L	100
21) Acrylonitrile	7.43	53	126153	113.2272 ug/L	98
22) Methyl Tert Butyl Ether	7.46	73	1894494	105.8643 ug/L	87
23) trans-1,2-Dichloroethene	7.70	61	1203701	99.1822 ug/L	93
24) n-Hexane	7.78	57	926226	105.3097 ug/L	84
25) Diisopropyl ether	8.12	45	3699807	196.0386 ug/L #	90
26) Vinyl Acetate	8.29	43	860820	101.8643 ug/L #	88
27) 1,1-Dichloroethane	8.32	63	1521237	101.5567 ug/L	96
28) Ethyl-Tert-Butyl ether	8.70	59	4005676	198.2271 ug/L	89
29) 2-Butanone	8.88	43	145126	112.6698 ug/L #	81
30) Propionitrile	8.98	54	83038	219.1648 ug/L	97
31) 2,2-Dichloropropane	9.11	77	1610615	102.0557 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	935490	99.9944 ug/L	89
33) Chloroform	9.37	83	1766988	98.8341 ug/L	98
34) 1-Bromopropane	9.51	122	139657	108.5299 ug/L	100
35) Bromochloromethane	9.61	130	488678	102.8480 ug/L	66
36) Tetrahydrofuran	9.63	42	167812	187.5699 ug/L #	61
38) 1,1,1-Trichloroethane	9.91	97	1724504	102.7713 ug/L	100
39) Cyclohexane	9.94	56	1185537	102.8628 ug/L #	66
40) 1,1-Dichloropropene	10.11	75	1293224	100.8377 ug/L	96
41) Tert-Amyl-Methyl ether	10.21	73	3634806	198.1264 ug/L	88
42) Carbon Tetrachloride	10.25	117	1534696	103.3979 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M412227.D 8260WT.M Fri May 13 17:14:33 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412227.D Vial: 10
Acq On : 13 May 2016 16:52 Operator: TMB
Sample : WG568561-09 100ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:14:31 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:12:43 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc	Unit		Qvalue
45)	1,2-Dichloroethane	10.42	62	1177876	103.7352	ug/L		85
46)	Benzene	10.46	78	3298928	95.8405	ug/L		87
47)	Trichloroethene	11.21	130	903490	101.9036	ug/L		97
48)	Methylcyclohexane	11.30	83	1389528	104.6454	ug/L		83
49)	1,2-Dichloropropane	11.43	63	755128	102.5297	ug/L		73
50)	Bromodichloromethane	11.73	83	1351612	104.6036	ug/L		99
51)	1,4-Dioxane	11.72	88	16303	473.9624	ug/L		68
52)	Dibromomethane	11.81	93	445130	101.6543	ug/L		99
53)	2-Chloroethyl Vinyl Ether	12.02	63	333977	105.4228	ug/L		92
54)	4-Methyl-2-Pentanone	12.05	58	138980	110.1494	ug/L	#	77
55)	cis-1,3-Dichloropropene	12.36	75	1387475	105.9136	ug/L		100
56)	Dimethyl Disulfide	12.62	79	745504	111.3169	ug/L		97
59)	Toluene	12.78	91	3455614	95.2346	ug/L		99
60)	Ethyl Methacrylate	12.88	69	731009	113.5760	ug/L		70
62)	trans-1,3-Dichloropropene	12.96	75	1236182	107.9445	ug/L		99
63)	1,1,2-Trichloroethane	13.18	97	564076	103.0604	ug/L		99
64)	2-Hexanone	13.12	58	126268	114.5366	ug/L	#	79
65)	1,3-Dichloropropane	13.48	76	1021082	105.2672	ug/L		74
66)	Tetrachloroethene	13.60	164	752835	101.4023	ug/L		95
67)	Dibromochloromethane	13.87	129	856499	109.6525	ug/L		100
68)	1,2-Dibromoethane	14.12	107	577701	107.6426	ug/L		98
69)	1-Chlorohexane	14.21	91	1222720	103.0385	ug/L		82
70)	Chlorobenzene	14.62	112	2332247	97.2408	ug/L		96
71)	1,1,1,2-Tetrachloroethane	14.66	131	947925	106.1463	ug/L		99
72)	Ethylbenzene	14.66	106	1275519	98.8771	ug/L		97
73)	m-,p-Xylene	14.75	106	3009424	194.2602	ug/L		98
74)	o-Xylene	15.31	106	1490638	100.8068	ug/L		93
75)	Styrene	15.35	104	2503721	102.8373	ug/L		88
76)	Bromoform	15.83	173	517827	114.0520	ug/L		100
77)	Isopropylbenzene	15.73	105	3913141	95.4433	ug/L		99
79)	1,1,2,2-Tetrachloroethane	15.95	83	624965	106.8610	ug/L		97
81)	1,2,3-Trichloropropane	16.14	110	192087	107.3268	ug/L		97
82)	trans-1,4-Dichloro-2-Buten	16.18	53	189214	128.4818	ug/L	#	59
83)	n-Propylbenzene	16.24	91	4599057	93.6644	ug/L		100
84)	Bromobenzene	16.37	156	968011	102.1426	ug/L		95
85)	1,3,5-Trimethylbenzene	16.43	105	3401719	97.9556	ug/L		97
86)	2-Chlorotoluene	16.52	91	3177047	93.8267	ug/L		98
87)	4-Chlorotoluene	16.56	91	2899171	96.0712	ug/L		96
88)	a-Methylstyrene	16.84	118	1639166	103.6811	ug/L		90
89)	tert-Butylbenzene	16.90	134	654025	99.4199	ug/L		85
90)	1,2,4-Trimethylbenzene	16.95	105	3479295	96.7054	ug/L		97
91)	sec-Butylbenzene	17.17	105	3947832	94.2785	ug/L		98
92)	p-Isopropyltoluene	17.32	119	3328450	94.4148	ug/L		97
93)	1,3-Dichlorobenzene	17.52	146	1834740	98.9074	ug/L		92
94)	1,4-Dichlorobenzene	17.64	146	1818309	96.6271	ug/L		92
95)	n-Butylbenzene	17.85	91	3308330	94.4002	ug/L		98
96)	1,2-Dichlorobenzene	18.14	146	1606254	97.0179	ug/L		91
97)	1,2-Dibromo-3-Chloropropan	19.13	75	125368	107.7020	ug/L		87
98)	1,2,4-Trichlorobenzene	20.28	180	1227926	100.7076	ug/L		99
99)	Hexachlorobutadiene	20.43	225	569213	100.5231	ug/L		97
100)	Naphthalene	20.65	128	1852486	101.2968	ug/L		99
101)	1,2,3-Trichlorobenzene	20.96	180	1021163	98.5794	ug/L		100

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412227.D Vial: 10
Acq On : 13 May 2016 16:52 Operator: TMB
Sample : WG568561-09 100ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:14 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:12:43 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412228.D Vial: 11
Acq On : 13 May 2016 17:21 Operator: TMB
Sample : WG568561-10 200ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:43:32 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	828540	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	578848	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	301401	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	898145	101.4517 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 405.80%#		
43) 1,2-Dichloroethane-d4	10.30	65	939848	98.0779 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 392.32%#		
58) Toluene-d8	12.68	98	2984162	97.8805 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 391.52%#		
80) p-Bromofluorobenzene	16.08	95	1242720	99.7023 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 398.80%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	2830517	203.5621 ug/L	98
3) Chloromethane	3.62	50	1586058	194.4056 ug/L	97
4) Vinyl Chloride	3.83	62	1935009	201.0738 ug/L	94
5) 1,3-Butadiene	3.88	54	1684964	212.5477 ug/L	94
6) Bromomethane	4.72	94	1358068	271.2231 ug/L	99
7) Chloroethane	4.89	64	1084803	206.5255 ug/L	90
8) Trichlorofluoromethane	5.37	101	3516760	195.1756 ug/L	99
10) Isoprene	5.94	67	2136665	196.1573 ug/L	78
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	1662142	196.5993 ug/L	95
13) Acetone	6.24	43	187748	200.7963 ug/L	97
14) 1,1-Dichloroethene	6.46	61	2649821	196.6732 ug/L	92
15) Tert-Butyl Alcohol	6.58	59	6024	20.0950 ug/L #	87
16) Dimethyl Sulfide	6.73	62	976862	205.8822 ug/L	82
17) Iodomethane	6.99	142	896947	398.7987 ug/L	100
18) Methyl acetate	7.00	43	504293	201.4324 ug/L #	81
19) Methylene Chloride	7.25	84	1684780	196.5546 ug/L	52
20) Carbon Disulfide	7.29	76	5386210	189.2120 ug/L	99
21) Acrylonitrile	7.45	53	30354	26.0447 ug/L #	32
22) Methyl Tert Butyl Ether	7.46	73	3711011	200.6529 ug/L	87
23) trans-1,2-Dichloroethene	7.70	61	2410498	194.2902 ug/L	92
24) n-Hexane	7.78	57	1821777	199.7288 ug/L	84
26) Vinyl Acetate	8.29	43	1796217	206.9060 ug/L #	88
27) 1,1-Dichloroethane	8.32	63	3042055	197.9863 ug/L	95
29) 2-Butanone	8.88	43	271767	199.8193 ug/L #	77
30) Propionitrile	8.90	54	208	0.5280 ug/L #	1
31) 2,2-Dichloropropane	9.11	77	3202633	197.6970 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	1875631	195.8883 ug/L	89
33) Chloroform	9.37	83	3506648	191.9199 ug/L	99
34) 1-Bromopropane	9.51	122	286158	214.2313 ug/L	99
35) Bromochloromethane	9.60	130	980373	200.7807 ug/L	66
36) Tetrahydrofuran	9.62	42	2381	2.6275 ug/L	87
38) 1,1,1-Trichloroethane	9.91	97	3410306	197.7909 ug/L	99
39) Cyclohexane	9.94	56	2357393	198.8974 ug/L #	66
40) 1,1-Dichloropropene	10.10	75	2601175	197.8946 ug/L	96
42) Carbon Tetrachloride	10.25	117	3062767	200.6417 ug/L	99
45) 1,2-Dichloroethane	10.42	62	2313662	198.0324 ug/L	85
46) Benzene	10.46	78	6379817	182.1771 ug/L	89
47) Trichloroethene	11.21	130	1810459	198.9741 ug/L	97
48) Methylcyclohexane	11.30	83	2755373	200.4193 ug/L	83
49) 1,2-Dichloropropane	11.43	63	1528407	202.0331 ug/L	74

(#) = qualifier out of range (m) = manual integration
8M412228.D 8260WT.M Fri May 13 17:43:34 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412228.D Vial: 11
Acq On : 13 May 2016 17:21 Operator: TMB
Sample : WG568561-10 200ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:43:32 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
50)	Bromodichloromethane	11.73	83	2716764	204.0895 ug/L		99
51)	1,4-Dioxane	11.73	88	3865	104.9352 ug/L	#	18
52)	Dibromomethane	11.81	93	898971	200.1154 ug/L		98
53)	2-Chloroethyl Vinyl Ether	12.02	63	631656	192.7236 ug/L		90
54)	4-Methyl-2-Pentanone	12.05	58	262702	198.3957 ug/L	#	75
55)	cis-1,3-Dichloropropene	12.36	75	2784162	205.9153 ug/L		99
56)	Dimethyl Disulfide	12.62	79	1518705	215.4711 ug/L		95
59)	Toluene	12.78	91	6564353	175.9274 ug/L		96
60)	Ethyl Methacrylate	12.88	69	1429837	209.8161 ug/L		71
62)	trans-1,3-Dichloropropene	12.96	75	2468609	205.4771 ug/L		98
63)	1,1,2-Trichloroethane	13.18	97	1124672	197.6019 ug/L		100
64)	2-Hexanone	13.12	58	237678	200.9290 ug/L	#	81
65)	1,3-Dichloropropane	13.48	76	2023247	199.9545 ug/L		76
66)	Tetrachloroethene	13.60	164	1546618	200.8018 ug/L		97
67)	Dibromochloromethane	13.87	129	1726523	210.5823 ug/L		100
68)	1,2-Dibromoethane	14.12	107	1163160	207.0667 ug/L		99
69)	1-Chlorohexane	14.21	91	2414147	195.6413 ug/L		83
70)	Chlorobenzene	14.62	112	4641169	187.6386 ug/L		94
71)	1,1,1,2-Tetrachloroethane	14.66	131	1929251	206.8372 ug/L		99
72)	Ethylbenzene	14.66	106	2611321	195.8267 ug/L		88
73)	m-,p-Xylene	14.75	106	5876234	367.8657 ug/L		83
74)	o-Xylene	15.31	106	3012318	196.4899 ug/L		98
75)	Styrene	15.35	104	4913204	194.1235 ug/L		86
76)	Bromoform	15.83	173	1052629	218.7989 ug/L		100
77)	Isopropylbenzene	15.73	105	7284624	172.7305 ug/L		97
79)	1,1,2,2-Tetrachloroethane	15.95	83	1237171	205.8673 ug/L		98
81)	1,2,3-Trichloropropane	16.14	110	386738	209.7906 ug/L	#	54
82)	trans-1,4-Dichloro-2-Buten	16.19	53	382467	243.6526 ug/L		76
83)	n-Propylbenzene	16.24	91	8244387	166.5110 ug/L		93
84)	Bromobenzene	16.37	156	1935166	200.1305 ug/L		95
85)	1,3,5-Trimethylbenzene	16.43	105	6482975	183.9948 ug/L		98
86)	2-Chlorotoluene	16.52	91	6108308	178.8549 ug/L		99
87)	4-Chlorotoluene	16.56	91	5513221	180.5508 ug/L		99
88)	a-Methylstyrene	16.84	118	3255254	200.4995 ug/L		89
89)	tert-Butylbenzene	16.90	134	1351569	202.1005 ug/L		92
90)	1,2,4-Trimethylbenzene	16.95	105	6531416	179.2445 ug/L		97
91)	sec-Butylbenzene	17.17	105	7342730	173.9814 ug/L		97
92)	p-Isopropyltoluene	17.32	119	6352010	178.7315 ug/L		99
93)	1,3-Dichlorobenzene	17.52	146	3640547	193.1654 ug/L		90
94)	1,4-Dichlorobenzene	17.64	146	3593548	188.4603 ug/L		91
95)	n-Butylbenzene	17.85	91	6198060	175.4375 ug/L		97
96)	1,2-Dichlorobenzene	18.14	146	3171699	188.9651 ug/L		89
97)	1,2-Dibromo-3-Chloropropan	19.13	75	249646	207.5645 ug/L		88
98)	1,2,4-Trichlorobenzene	20.28	180	2480626	199.7296 ug/L		99
99)	Hexachlorobutadiene	20.43	225	1169411	202.7981 ug/L		98
100)	Naphthalene	20.65	128	3580198	192.0323 ug/L		97
101)	1,2,3-Trichlorobenzene	20.96	180	2040570	193.9298 ug/L		100

(#) = qualifier out of range (m) = manual integration
8M412228.D 8260WT.M Fri May 13 17:43:35 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412228.D Vial: 11
Acq On : 13 May 2016 17:21 Operator: TMB
Sample : WG568561-10 200ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:43 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412229.D Vial: 12
Acq On : 13 May 2016 17:50 Operator: TMB
Sample : WG568561-11 300ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 18:12:33 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	855409	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	599314	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	305726	25.00 ug/L	0.00

System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	1411837	154.4674 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 617.88%#		
43) 1,2-Dichloroethane-d4	10.30	65	1466019	148.1811 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 592.72%#		
58) Toluene-d8	12.69	98	4594713	145.5601 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 582.24%#		
80) p-Bromofluorobenzene	16.08	95	1918996	151.7813 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 607.12%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	4191932	292.0014 ug/L	97
3) Chloromethane	3.62	50	2474035	293.7212 ug/L	97
4) Vinyl Chloride	3.83	62	2956347	297.5552 ug/L	94
5) 1,3-Butadiene	3.87	54	2605506	318.3447 ug/L	94
6) Bromomethane	4.72	94	2177550	421.2239 ug/L	98
7) Chloroethane	4.89	64	1688005	311.2693 ug/L	90
8) Trichlorofluoromethane	5.37	101	5387662	289.6161 ug/L	99
9) Diethyl ether	5.90	59	1306432	295.4610 ug/L	80
10) Isoprene	5.93	67	3377164	300.3032 ug/L	78
11) Acrolein	6.14	56	70572	185.7401 ug/L	97
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	2563878	293.7317 ug/L	94
13) Acetone	6.24	43	283185	293.3529 ug/L	90
14) 1,1-Dichloroethene	6.46	61	4031745	289.8421 ug/L	91
15) Tert-Butyl Alcohol	6.58	59	181796	587.3907 ug/L	90
16) Dimethyl Sulfide	6.73	62	1553493	317.1280 ug/L	81
17) Iodomethane	6.99	142	1347307	580.2206 ug/L	98
18) Methyl acetate	7.00	43	800046	309.5287 ug/L #	80
19) Methylene Chloride	7.24	84	2587731	292.4144 ug/L	52
20) Carbon Disulfide	7.29	76	8228181	279.9683 ug/L	98
21) Acrylonitrile	7.43	53	228227	189.6753 ug/L	94
22) Methyl Tert Butyl Ether	7.46	73	5605924	293.5892 ug/L	88
23) trans-1,2-Dichloroethene	7.70	61	3678699	287.1957 ug/L	92
24) n-Hexane	7.78	57	2863434	304.0691 ug/L	86
25) Diisopropyl ether	8.12	45	5819647	292.7898 ug/L #	89
26) Vinyl Acetate	8.30	43	2819503	314.5767 ug/L #	87
27) 1,1-Dichloroethane	8.32	63	4586531	289.1293 ug/L	94
28) Ethyl-Tert-Butyl ether	8.70	59	6190206	290.3321 ug/L	89
29) 2-Butanone	8.88	43	439359	312.8959 ug/L #	76
30) Propionitrile	8.99	54	133197	327.4673 ug/L	95
31) 2,2-Dichloropropane	9.10	77	4821388	288.2735 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	2847005	287.9976 ug/L	89
33) Chloroform	9.37	83	5197860	275.5447 ug/L	100
34) 1-Bromopropane	9.51	122	450012	326.3179 ug/L	100
35) Bromochloromethane	9.61	130	1490221	295.6112 ug/L	66
36) Tetrahydrofuran	9.63	42	271736	290.4491 ug/L #	62
38) 1,1,1-Trichloroethane	9.91	97	5126354	287.9792 ug/L	98
39) Cyclohexane	9.94	56	3733392	305.0987 ug/L #	67
40) 1,1-Dichloropropene	10.11	75	3923871	289.1471 ug/L	97
41) Tert-Amyl-Methyl ether	10.21	73	5639573	291.3713 ug/L	88
42) Carbon Tetrachloride	10.25	117	4598150	291.7629 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M412229.D 8260WT.M Fri May 13 18:12:35 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412229.D Vial: 12
Acq On : 13 May 2016 17:50 Operator: TMB
Sample : WG568561-11 300ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 18:12:33 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit	Qvalue
45)	1,2-Dichloroethane	10.43	62	3468738	287.5726 ug/L	86
46)	Benzene	10.46	78	9065861	250.7462 ug/L	93
47)	Trichloroethene	11.21	130	2772282	295.1106 ug/L	97
48)	Methylcyclohexane	11.30	83	4327357	304.8748 ug/L	84
49)	1,2-Dichloropropane	11.43	63	2350729	300.9718 ug/L	74
50)	Bromodichloromethane	11.73	83	4051493	294.7973 ug/L	99
51)	1,4-Dioxane	11.72	88	26614	699.8767 ug/L	67
52)	Dibromomethane	11.81	93	1369945	295.3776 ug/L	97
53)	2-Chloroethyl Vinyl Ether	12.03	63	1006317	297.3917 ug/L	90
54)	4-Methyl-2-Pentanone	12.06	58	424684	310.6522 ug/L #	76
55)	cis-1,3-Dichloropropene	12.36	75	4156310	297.7431 ug/L	97
56)	Dimethyl Disulfide	12.63	79	2359825	324.2911 ug/L	94
59)	Toluene	12.78	91	9135213	236.4669 ug/L	90
60)	Ethyl Methacrylate	12.88	69	2228037	315.7801 ug/L	71
62)	trans-1,3-Dichloropropene	12.96	75	3700845	297.5240 ug/L	96
63)	1,1,2-Trichloroethane	13.18	97	1730199	293.6104 ug/L	100
64)	2-Hexanone	13.11	58	381967	311.8816 ug/L #	78
65)	1,3-Dichloropropane	13.49	76	3080177	294.0141 ug/L	76
66)	Tetrachloroethene	13.60	164	2374435	297.7522 ug/L	96
67)	Dibromochloromethane	13.87	129	2638145	310.7838 ug/L	100
68)	1,2-Dibromoethane	14.13	107	1778561	305.8085 ug/L	99
69)	1-Chlorohexane	14.21	91	3712592	290.5924 ug/L	84
70)	Chlorobenzene	14.63	112	6754240	263.7433 ug/L	90
71)	1,1,1,2-Tetrachloroethane	14.65	131	2967650	307.3001 ug/L	100
72)	Ethylbenzene	14.65	106	4014412	290.7661 ug/L	70
73)	m-,p-Xylene	14.75	106	8357822	505.3514 ug/L	70
74)	o-Xylene	15.31	106	4530840	285.4488 ug/L	88
75)	Styrene	15.35	104	7050702	269.0641 ug/L	82
76)	Bromoform	15.84	173	1602587	321.7373 ug/L	99
77)	Isopropylbenzene	15.73	105	9900817	226.7477 ug/L	90
79)	1,1,2,2-Tetrachloroethane	15.95	83	1885953	309.3861 ug/L	98
81)	1,2,3-Trichloropropane	16.14	110	581198	310.8176 ug/L	61
82)	trans-1,4-Dichloro-2-Buten	16.18	53	589705	370.3601 ug/L	82
83)	n-Propylbenzene	16.25	91	10768378	214.4110 ug/L	86
84)	Bromobenzene	16.37	156	2947671	300.5290 ug/L	99
85)	1,3,5-Trimethylbenzene	16.43	105	8890651	248.7580 ug/L	92
86)	2-Chlorotoluene	16.51	91	7868844	227.1449 ug/L	92
87)	4-Chlorotoluene	16.57	91	7782731	251.2687 ug/L	93
88)	a-Methylstyrene	16.84	118	4955768	300.9205 ug/L	86
89)	tert-Butylbenzene	16.90	134	2046401	301.6701 ug/L	95
90)	1,2,4-Trimethylbenzene	16.95	105	8903769	240.8932 ug/L	89
91)	sec-Butylbenzene	17.17	105	9867862	230.5051 ug/L	91
92)	p-Isopropyltoluene	17.32	119	8738927	242.4155 ug/L	94
93)	1,3-Dichlorobenzene	17.52	146	5319954	278.2806 ug/L	87
94)	1,4-Dichlorobenzene	17.64	146	5248511	271.3593 ug/L	88
95)	n-Butylbenzene	17.85	91	8408157	234.6279 ug/L	92
96)	1,2-Dichlorobenzene	18.14	146	4717385	277.0787 ug/L	87
97)	1,2-Dibromo-3-Chloropropan	19.13	75	377070	309.0741 ug/L	91
98)	1,2,4-Trichlorobenzene	20.28	180	3677657	291.9206 ug/L	98
99)	Hexachlorobutadiene	20.43	225	1780695	304.4378 ug/L	96
100)	Naphthalene	20.65	128	5216751	275.8542 ug/L	95
101)	1,2,3-Trichlorobenzene	20.96	180	3064908	287.1593 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M412229.D 8260WT.M Fri May 13 18:12:36 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412229.D Vial: 12
Acq On : 13 May 2016 17:50 Operator: TMB
Sample : WG568561-11 300ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 18:12 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration





Method Name: K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:08:06 2016



Method Name: K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:08:06 2016



Method Name: K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:08:06 2016

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\2\DATA\051316\8M412232.D Vial: 15
Acq On : 13 May 2016 19:17 Operator: TMB
Sample : WG568561-12 50ug/L ALT SRC STD 8260 Inst : HPMS8
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 17 11:44:36 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Tue May 17 11:43:20 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	837879	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	574720	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	304324	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.67	111	229742	25.5204 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 102.08%		
43) 1,2-Dichloroethane-d4	10.31	65	238672	24.7259 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 98.92%		
58) Toluene-d8	12.68	98	776362	25.8115 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 103.24%		
80) p-Bromofluorobenzene	16.07	95	315683	25.0519 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 100.20%		
Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.17	85	761591	54.2141 ug/L	99
3) Chloromethane	3.62	50	415589	50.6814 ug/L	100
4) Vinyl Chloride	3.84	62	509607	52.3810 ug/L	100
5) 1,3-Butadiene	3.88	54	347562	42.6002 ug/L	97
6) Bromomethane	4.73	94	272060	46.3924 ug/L	99
7) Chloroethane	4.88	64	285509	53.3336 ug/L	98
8) Trichlorofluoromethane	5.37	101	849580	46.9313 ug/L	100
9) Diethyl ether	5.90	59	451307	104.4282 ug/L	99
10) Isoprene	5.93	67	582863	53.0208 ug/L	99
11) Acrolein	6.13	56	38877	100.4722 ug/L	99
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	419584	49.3091 ug/L	97
13) Acetone	6.23	43	54234	57.5310 ug/L	94
14) 1,1-Dichloroethene	6.46	61	622831	45.9700 ug/L	99
15) Tert-Butyl Alcohol	6.59	59	59797	197.9422 ug/L	97
16) Dimethyl Sulfide	6.73	62	351814	72.6234 ug/L	98
17) Iodomethane	6.99	142	92716	26.2832 ug/L	99
18) Methyl acetate	7.00	43	127540	50.0514 ug/L	98
19) Methylene Chloride	7.25	84	417642	48.4385 ug/L	99
20) Carbon Disulfide	7.29	76	1199899	42.3200 ug/L	100
21) Acrylonitrile	7.43	53	65324	53.4074 ug/L	93
22) Methyl Tert Butyl Ether	7.46	73	949572	50.8859 ug/L	97
23) trans-1,2-Dichloroethene	7.70	61	598296	48.0666 ug/L	99
24) n-Hexane	7.77	57	425106	45.9930 ug/L	99
25) Diisopropyl ether	8.13	45	2040382	105.1615 ug/L	98
26) Vinyl Acetate	8.29	43	415568	46.7802 ug/L	99
27) 1,1-Dichloroethane	8.32	63	721661	46.6846 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	2040020	98.1343 ug/L	99
29) 2-Butanone	8.88	43	73394	52.9905 ug/L	99
30) Propionitrile	8.98	54	42290	104.7756 ug/L	100
31) 2,2-Dichloropropane	9.10	77	780073	47.8861 ug/L	99
32) cis-1,2-Dichloroethene	9.16	96	470465	48.9162 ug/L	98
33) Chloroform	9.38	83	860252	47.1317 ug/L	100
34) 1-Bromopropane	9.51	122	98585	71.5615 ug/L	98
35) Bromochloromethane	9.60	130	234493	47.5456 ug/L	99
36) Tetrahydrofuran	9.62	42	86049	94.3281 ug/L	99
38) 1,1,1-Trichloroethane	9.90	97	825167	47.5949 ug/L	100
39) Cyclohexane	9.93	56	586721	48.8808 ug/L	97
40) 1,1-Dichloropropene	10.10	75	627217	47.4632 ug/L	99
41) Tert-Amyl-Methyl ether	10.21	73	1962875	103.9618 ug/L	99
42) Carbon Tetrachloride	10.24	117	711823	46.2363 ug/L	100

(#) = qualifier out of range (m) = manual integration
8M412232.D 8260WT.M Tue May 17 11:44:38 2016 Page 1

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412232.D Vial: 15
Acq On : 13 May 2016 19:17 Operator: TMB
Sample : WG568561-12 50ug/L ALT SRC STD 8260 Inst : HPMS8
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	108	0.00
2	T	Dichlorodifluoromethane	50.0000	54.2141	-8.4	112	0.00
3	P	Chloromethane	50.0000	50.6814	-1.4	111	0.00
4	C	Vinyl Chloride	50.0000	52.3811	-4.8	112	-0.01
5	T	1,3-Butadiene	50.0000	42.6002	14.8	89	0.00
6	T	Bromomethane	50.0000	46.3924	7.2	104	-0.01
7	T	Chloroethane	50.0000	53.3336	-6.7	117	-0.01
8	T	Trichlorofluoromethane	50.0000	46.9313	6.1	101	-0.01
9	T	Diethyl ether	100.0000	104.4282	-4.4	116	0.00
10	T	Isoprene	50.0000	53.0208	-6.0	117	0.00
11	T	Acrolein	50.0000	100.4722	-100.9#	220	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	50.0000	49.3091	1.4	108	0.00
13	T	Acetone	50.0000	57.5310	-15.1	118	-0.01
14	C	1,1-Dichloroethene	50.0000	45.9700	8.1	100	0.00
15	T	Tert-Butyl Alcohol	200.0000	197.9423	1.0	108	0.01
16	T	Dimethyl Sulfide	50.0000	72.6234	-45.2#	159	0.00
17	T	Iodomethane	50.0000	26.2832	47.4#	59	0.01
18	T	Methyl acetate	50.0000	50.0514	-0.1	111	0.00
19	T	Methylene Chloride	50.0000	48.4385	3.1	107	0.00
20	T	Carbon Disulfide	50.0000	42.3200	15.4	94	0.00
21	T	Acrylonitrile	50.0000	53.4074	-6.8	109	0.00
22	T	Methyl Tert Butyl Ether	50.0000	50.8859	-1.8	105	0.00
23	T	trans-1,2-Dichloroethene	50.0000	48.0666	3.9	105	0.00
24	T	n-Hexane	50.0000	45.9930	8.0	101	-0.01
25	T	Diisopropyl ether	100.0000	105.1615	-5.2	115	0.00
26	T	Vinyl Acetate	50.0000	46.7802	6.4	97	0.00
27	P	1,1-Dichloroethane	50.0000	46.6846	6.6	100	0.00
28	T	Ethyl-Tert-Butyl ether	100.0000	98.1343	1.9	104	0.00
29	T	2-Butanone	50.0000	52.9905	-6.0	110	0.00
30	T	Propionitrile	100.0000	104.7756	-4.8	111	0.00
31	T	2,2-Dichloropropane	50.0000	47.8861	4.2	102	0.00
32	T	cis-1,2-Dichloroethene	50.0000	48.9162	2.2	107	-0.01
33	C	Chloroform	50.0000	47.1317	5.7	102	0.00
34		1-Bromopropane	50.0000	71.5615	-43.1#	155	0.00
35	T	Bromochloromethane	50.0000	47.5456	4.9	102	0.00
36	T	Tetrahydrofuran	100.0000	94.3281	5.7	108	-0.01
37	S	Dibromofluoromethane	25.0000	25.5204	-2.1	109	0.00
38	T	1,1,1-Trichloroethane	50.0000	47.5949	4.8	101	0.00
39	T	Cyclohexane	50.0000	48.8808	2.2	109	0.00
40	T	1,1-Dichloropropene	50.0000	47.4632	5.1	102	0.00
41	T	Tert-Amyl-Methyl ether	100.0000	103.9618	-4.0	111	0.00
42	T	Carbon Tetrachloride	50.0000	46.2363	7.5	98	0.00
43	S	1,2-Dichloroethane-d4	25.0000	24.7259	1.1	104	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	50.0000	49.0148	2.0	103	0.00
46	T	Benzene	50.0000	48.5232	3.0	104	0.00
47	T	Trichloroethene	50.0000	51.0658	-2.1	110	0.00
48	T	Methylcyclohexane	50.0000	49.8573	0.3	109	0.00
49	C	1,2-Dichloropropane	50.0000	50.9747	-1.9	111	0.00
50	T	Bromodichloromethane	50.0000	48.3053	3.4	99	0.00
51	T	1,4-Dioxane	200.0000	216.9538	-8.5	110	0.01
52	T	Dibromomethane	50.0000	45.8126	8.4	99	0.00
53	T	2-Chloroethyl Vinyl Ether	50.0000	51.2623	-2.5	104	0.00
54	T	4-Methyl-2-Pentanone	50.0000	51.8715	-3.7	107	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412232.D Vial: 15
Acq On : 13 May 2016 19:17 Operator: TMB
Sample : WG568561-12 50ug/L ALT SRC STD 8260 Inst : HPMS8
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	50.0000	53.8979	-7.8	111	0.00
56	T	Dimethyl Disulfide	50.0000	52.2714	-4.5	111	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	107	0.00
58	S	Toluene-d8	25.0000	25.8115	-3.2	108	0.00
59	C	Toluene	50.0000	49.0912	1.8	105	0.00
60	T	Ethyl Methacrylate	50.0000	55.9509	-11.9	113	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	50.0000	50.6740	-1.3	102	0.00
63	T	1,1,2-Trichloroethane	50.0000	49.0377	1.9	103	0.00
64	T	2-Hexanone	50.0000	53.6057	-7.2	112	0.00
65	T	1,3-Dichloropropane	50.0000	53.6720	-7.3	110	0.00
66	T	Tetrachloroethene	50.0000	47.8053	4.4	104	0.00
67	T	Dibromochloromethane	50.0000	49.6585	0.7	102	0.00
68	T	1,2-Dibromoethane	50.0000	51.1121	-2.2	105	0.00
69	T	1-Chlorohexane	50.0000	51.0611	-2.1	111	0.00
70	P	Chlorobenzene	50.0000	50.1844	-0.4	108	0.00
71	T	1,1,1,2-Tetrachloroethane	50.0000	49.3793	1.2	103	0.00
72	C	Ethylbenzene	50.0000	48.2013	3.6	105	0.00
73	T	m-,p-Xylene	100.0000	101.1341	-1.1	108	0.00
74	T	o-Xylene	50.0000	50.5675	-1.1	108	0.00
75	T	Styrene	50.0000	51.5086	-3.0	105	0.00
76	P	Bromoform	50.0000	48.2227	3.6	99	0.00
77	T	Isopropylbenzene	50.0000	51.1140	-2.2	107	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	106	0.00
79	P	1,1,2,2-Tetrachloroethane	50.0000	49.6172	0.8	103	0.00
80	S	p-Bromofluorobenzene	25.0000	25.0519	-0.2	108	0.00
81	T	1,2,3-Trichloropropane	50.0000	50.0849	-0.2	106	0.01
82	T	trans-1,4-Dichloro-2-Butene	50.0000	39.2043	21.6	86	0.00
83	T	n-Propylbenzene	50.0000	51.9237	-3.8	108	0.00
84	T	Bromobenzene	50.0000	49.4648	1.1	105	0.00
85	T	1,3,5-Trimethylbenzene	50.0000	51.6344	-3.3	107	0.00
86	T	2-Chlorotoluene	50.0000	51.5639	-3.1	106	0.00
87	T	4-Chlorotoluene	50.0000	48.2650	3.5	108	0.00
88	T	a-Methylstyrene	50.0000	56.4132	-12.8	120	0.00
89	T	tert-Butylbenzene	50.0000	50.9494	-1.9	112	0.00
90	T	1,2,4-Trimethylbenzene	50.0000	49.6617	0.7	105	0.00
91	T	sec-Butylbenzene	50.0000	51.3386	-2.7	108	0.00
92	T	p-Isopropyltoluene	50.0000	50.1261	-0.3	107	0.00
93	T	1,3-Dichlorobenzene	50.0000	50.8943	-1.8	108	0.00
94	T	1,4-Dichlorobenzene	50.0000	50.4742	-0.9	110	0.00
95	T	n-Butylbenzene	50.0000	49.6520	0.7	105	0.00
96	T	1,2-Dichlorobenzene	50.0000	50.4907	-1.0	110	0.00
97	T	1,2-Dibromo-3-Chloropropane	50.0000	48.4851	3.0	100	0.00
98	T	1,2,4-Trichlorobenzene	50.0000	50.1129	-0.2	107	0.00
99	T	Hexachlorobutadiene	50.0000	52.0919	-4.2	115	0.00
100	T	Naphthalene	50.0000	45.8047	8.4	95	0.00
101	T	1,2,3-Trichlorobenzene	50.0000	47.7353	4.5	106	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412232.D 8260WT.M Sat May 14 18:09:03 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412321.D Vial: 3
Acq On : 17 May 2016 9:47 Operator: TMB
Sample : WG569079-01 100ug/L A9 CCV STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 09:01:21 2016 Quant Results File: A9FOOWT.RES

Quant Method : K:\ORGANICS\V...\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	709364	25.00 ug/L	0.00
11) Chlorobenzene-d5	14.57	117	505399	25.00 ug/L	0.00
12) 1,4-Dichlorobenzene-d4	17.60	152	272922	25.00 ug/L	0.00

Target Compounds					Qvalue
2) Acetonitrile	6.69	41	27428	100.2132 ug/L	99
3) 3-Chloro-1-propene	7.07	41	940418	104.3259 ug/L	82
4) 2-Chloro-1,3-butadiene	8.45	53	1188994	104.3183 ug/L	86
5) Ethyl Acetate	9.10	43	283167	104.3264 ug/L	95
6) Methacrylonitrile	9.25	67	155775	104.5365 ug/L #	63
7) Isobutyl Alcohol	9.25	43	17893	197.3629 ug/L #	96
9) Methyl methacrylate	11.42	41	298802	102.0346 ug/L	70
10) 2-Nitropropane	11.75	43	121321	95.8263 ug/L	97
13) Cyclohexanone	15.84	55	11401	102.1363 ug/L #	70

(#) = qualifier out of range (m) = manual integration
8M412321.D A9FOOWT.M Thu May 19 09:01:22 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412321.D Vial: 3
Acq On : 17 May 2016 9:47 Operator: TMB
Sample : WG569079-01 100ug/L A9 CCV STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 9:00 2016 Quant Results File: A9FOOWT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412321.D Vial: 3
Acq On : 17 May 2016 9:47 Operator: TMB
Sample : WG569079-01 100ug/L A9 CCV STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 1% Max. R.T. Dev 0.50min
Max. RRF Dev : 75% Max. Rel. Area : 200%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Acetonitrile	100.0000	100.2132	-0.2	100	0.00
3	T	3-Chloro-1-propene	100.0000	104.3259	-4.3	100	0.00
4	T	2-Chloro-1,3-butadiene	100.0000	104.3183	-4.3	100	0.00
5	T	Ethyl Acetate	100.0000	104.3264	-4.3	100	0.00
6	T	Methacrylonitrile	100.0000	104.5365	-4.5	100	0.00
7	T	Isobutyl Alcohol	200.0000	197.3629	1.3	100	-0.01
8	T	1-Butanol	-1.0000	0.0000	0.0	0	0.02
9	T	Methyl methacrylate	100.0000	102.0346	-2.0	100	0.00
10	T	2-Nitropropane	100.0000	95.8263	4.2	100	0.00
11	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
12	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
13	T	Cyclohexanone	100.0000	102.1363	-2.1	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412321.D A9FOOWT.M Thu May 19 09:02:48 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412322.D Vial: 4
Acq On : 17 May 2016 10:16 Operator: TMB
Sample : WG569079-03 50ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 09:20:35 2016 Quant Results File: A9FOOWT.RES

Quant Method : K:\ORGANICS\V...\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	677339	25.00 ug/L	0.00
11) Chlorobenzene-d5	14.57	117	488377	25.00 ug/L	0.00
12) 1,4-Dichlorobenzene-d4	17.60	152	260939	25.00 ug/L	0.00

Target Compounds					Qvalue
2) Acetonitrile	6.69	41	13323	50.9795 ug/L	98
3) 3-Chloro-1-propene	7.07	41	453655	52.7060 ug/L	83
4) 2-Chloro-1,3-butadiene	8.45	53	568765	52.2609 ug/L	87
5) Ethyl Acetate	9.09	43	131636	50.7913 ug/L	97
6) Methacrylonitrile	9.25	67	74064	52.0523 ug/L #	62
7) Isobutyl Alcohol	9.26	43	8483	97.9930 ug/L	95
9) Methyl methacrylate	11.42	41	143892	51.4593 ug/L	70
10) 2-Nitropropane	11.75	43	56559	46.7858 ug/L	95
13) Cyclohexanone	15.84	55	5225	48.9579 ug/L	83

(#) = qualifier out of range (m) = manual integration
8M412322.D A9FOOWT.M Thu May 19 09:20:36 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412322.D Vial: 4
Acq On : 17 May 2016 10:16 Operator: TMB
Sample : WG569079-03 50ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 9:19 2016 Quant Results File: A9FOOWT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration



8M412322.D A9FOOWT.M Thu May 19 09:20:36 2016 Page 2

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412322.D Vial: 4
Acq On : 17 May 2016 10:16 Operator: TMB
Sample : WG569079-03 50ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 1% Max. R.T. Dev 0.50min
Max. RRF Dev : 75% Max. Rel. Area : 200%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Acetonitrile	50.0000	50.9795	-2.0	100	0.00
3	T	3-Chloro-1-propene	50.0000	52.7060	-5.4	100	0.00
4	T	2-Chloro-1,3-butadiene	50.0000	52.2609	-4.5	100	0.00
5	T	Ethyl Acetate	50.0000	50.7913	-1.6	100	0.00
6	T	Methacrylonitrile	50.0000	52.0523	-4.1	100	0.00
7	T	Isobutyl Alcohol	100.0000	97.9929	2.0	100	0.00
8	T	1-Butanol	-1.0000	0.0000	0.0	0	0.02
9	T	Methyl methacrylate	50.0000	51.4592	-2.9	100	0.00
10	T	2-Nitropropane	50.0000	46.7858	6.4	100	0.00
11	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
12	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
13	T	Cyclohexanone	50.0000	48.9579	2.1	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412322.D A9FOOWT.M Thu May 19 09:02:32 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412323.D Vial: 5
Acq On : 17 May 2016 10:44 Operator: TMB
Sample : WG569079-04 20ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 09:22:33 2016 Quant Results File: A9FOOWT.RES

Quant Method : K:\ORGANICS\V...\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	652985	25.00 ug/L	0.00
11) Chlorobenzene-d5	14.57	117	476539	25.00 ug/L	0.00
12) 1,4-Dichlorobenzene-d4	17.60	152	256082	25.00 ug/L	0.00

Target Compounds					Qvalue
2) Acetonitrile	6.69	41	4873	19.3416 ug/L	80
3) 3-Chloro-1-propene	7.07	41	172467	20.7847 ug/L	83
4) 2-Chloro-1,3-butadiene	8.45	53	219296	20.9015 ug/L	87
5) Ethyl Acetate	9.09	43	50385	20.1659 ug/L	100
6) Methacrylonitrile	9.26	67	28119	20.4992 ug/L #	63
7) Isobutyl Alcohol	9.27	43	2774	33.2395 ug/L #	8
9) Methyl methacrylate	11.42	41	55309	20.5176 ug/L	73
10) 2-Nitropropane	11.75	43	21584	18.5202 ug/L	99
13) Cyclohexanone	15.84	55	2230	21.2913 ug/L	87

(#) = qualifier out of range (m) = manual integration
8M412323.D A9FOOWT.M Thu May 19 09:22:33 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412323.D Vial: 5
Acq On : 17 May 2016 10:44 Operator: TMB
Sample : WG569079-04 20ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 9:21 2016 Quant Results File: A9FOOWT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration



8M412323.D A9FOOWT.M Thu May 19 09:22:34 2016 Page 2

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412323.D Vial: 5
Acq On : 17 May 2016 10:44 Operator: TMB
Sample : WG569079-04 20ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 1% Max. R.T. Dev 0.50min
Max. RRF Dev : 75% Max. Rel. Area : 200%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Acetonitrile	20.0000	19.3416	3.3	100	0.00
3	T	3-Chloro-1-propene	20.0000	20.7847	-3.9	100	0.00
4	T	2-Chloro-1,3-butadiene	20.0000	20.9015	-4.5	100	0.00
5	T	Ethyl Acetate	20.0000	20.1659	-0.8	100	0.00
6	T	Methacrylonitrile	20.0000	20.4992	-2.5	100	0.00
7	T	Isobutyl Alcohol	-1.0000	33.2395	0.0	0	0.00
8	T	1-Butanol	-1.0000	0.0000	0.0	0	-10.18#
9	T	Methyl methacrylate	20.0000	20.5176	-2.6	100	0.00
10	T	2-Nitropropane	20.0000	18.5203	7.4	100	0.00
11	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
12	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
13	T	Cyclohexanone	20.0000	21.2913	-6.5	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412323.D A9FOOWT.M Thu May 19 09:02:17 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412324.D Vial: 6
Acq On : 17 May 2016 11:13 Operator: TMB
Sample : WG569079-05 5ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 09:22:43 2016 Quant Results File: A9FOOWT.RES

Quant Method : K:\ORGANICS\V...\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	665645	25.00 ug/L	0.00
11) Chlorobenzene-d5	14.58	117	483812	25.00 ug/L	0.00
12) 1,4-Dichlorobenzene-d4	17.60	152	257678	25.00 ug/L	0.00

Target Compounds					Qvalue
2) Acetonitrile	6.70	41	405	1.5769 ug/L #	23
3) 3-Chloro-1-propene	7.07	41	40699	4.8115 ug/L	84
4) 2-Chloro-1,3-butadiene	8.45	53	48973	4.5789 ug/L	81
5) Ethyl Acetate	9.09	43	10675	4.1913 ug/L	95
6) Methacrylonitrile	9.26	67	5417	3.8740 ug/L #	72
7) Isobutyl Alcohol	9.27	43	441	5.1838 ug/L #	8
9) Methyl methacrylate	11.43	41	11811	4.2981 ug/L	65
10) 2-Nitropropane	11.76	43	4170	3.5100 ug/L	98

(#) = qualifier out of range (m) = manual integration
8M412324.D A9FOOWT.M Thu May 19 09:22:43 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412324.D Vial: 6
Acq On : 17 May 2016 11:13 Operator: TMB
Sample : WG569079-05 5ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 9:21 2016 Quant Results File: A9FOOWT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration



8M412324.D A9FOOWT.M Thu May 19 09:22:43 2016 Page 2

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412324.D Vial: 6
Acq On : 17 May 2016 11:13 Operator: TMB
Sample : WG569079-05 5ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 1% Max. R.T. Dev 0.50min
Max. RRF Dev : 75% Max. Rel. Area : 200%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Acetonitrile	-1.0000	1.5769	0.0	0	0.00
3	T	3-Chloro-1-propene	5.0000	4.8115	3.8	100	0.00
4	T	2-Chloro-1,3-butadiene	5.0000	4.5789	8.4	100	0.00
5	T	Ethyl Acetate	5.0000	4.1913	16.2	100	0.00
6	T	Methacrylonitrile	5.0000	3.8740	22.5	100	0.00
7	T	Isobutyl Alcohol	-1.0000	5.1838	0.0	0	0.00
8	T	1-Butanol	-1.0000	0.0000	0.0	0	-10.18#
9	T	Methyl methacrylate	5.0000	4.2981	14.0	100	0.00
10	T	2-Nitropropane	-1.0000	3.5100	0.0	0	0.00
11	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
12	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
13	T	Cyclohexanone	-1.0000	0.0000	0.0	0	-15.84#

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412324.D A9FOOWT.M Thu May 19 09:02:03 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412325.D Vial: 7
Acq On : 17 May 2016 11:42 Operator: TMB
Sample : WG569079-06 200ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 09:01:27 2016 Quant Results File: A9FOOWT.RES

Quant Method : K:\ORGANICS\V...\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	682012	25.00 ug/L	0.00
11) Chlorobenzene-d5	14.58	117	491515	25.00 ug/L	0.00
12) 1,4-Dichlorobenzene-d4	17.60	152	261877	25.00 ug/L	0.00

Target Compounds					Qvalue
2) Acetonitrile	6.68	41	52022	197.6946 ug/L	97
3) 3-Chloro-1-propene	7.06	41	1768403	204.0468 ug/L	83
4) 2-Chloro-1,3-butadiene	8.46	53	2283706	208.4003 ug/L	87
5) Ethyl Acetate	9.09	43	537585	206.0040 ug/L	94
6) Methacrylonitrile	9.25	67	297408	207.5869 ug/L #	63
7) Isobutyl Alcohol	9.26	43	34441	395.1256 ug/L	95
9) Methyl methacrylate	11.42	41	577612	205.1526 ug/L	72
10) 2-Nitropropane	11.75	43	248395	204.0652 ug/L	95
13) Cyclohexanone	15.85	55	19752	184.4122 ug/L #	59

(#) = qualifier out of range (m) = manual integration
8M412325.D A9FOOWT.M Thu May 19 09:01:28 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412325.D Vial: 7
Acq On : 17 May 2016 11:42 Operator: TMB
Sample : WG569079-06 200ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 9:00 2016 Quant Results File: A9FOOWT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412326.D Vial: 8
Acq On : 17 May 2016 12:40 Operator: TMB
Sample : WG569079-07 300ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 09:01:29 2016 Quant Results File: A9FOOWT.RES

Quant Method : K:\ORGANICS\V...\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	10.71	96	694903	25.00	ug/L	0.00
11) Chlorobenzene-d5	14.57	117	496643	25.00	ug/L	0.00
12) 1,4-Dichlorobenzene-d4	17.60	152	267309	25.00	ug/L	0.00

Target Compounds						Qvalue
2) Acetonitrile	6.69	41	81471	303.8636	ug/L	98
3) 3-Chloro-1-propene	7.06	41	2563559	290.3084	ug/L	83
4) 2-Chloro-1,3-butadiene	8.46	53	3270971	292.9562	ug/L	86
5) Ethyl Acetate	9.09	43	821184	308.8424	ug/L	95
6) Methacrylonitrile	9.25	67	451708	309.4374	ug/L #	62
7) Isobutyl Alcohol	9.26	43	55071	620.0836	ug/L	94
9) Methyl methacrylate	11.42	41	874430	304.8132	ug/L	70
10) 2-Nitropropane	11.75	43	392965	316.8456	ug/L	94
13) Cyclohexanone	15.84	55	33523	306.6233	ug/L #	61

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412326.D Vial: 8
Acq On : 17 May 2016 12:40 Operator: TMB
Sample : WG569079-07 300ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 9:00 2016 Quant Results File: A9FOOWT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412327.D Vial: 9
Acq On : 17 May 2016 13:10 Operator: TMB
Sample : WG569079-08 400ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 09:01:30 2016 Quant Results File: A9FOOWT.RES

Quant Method : K:\ORGANICS\V...\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	700514	25.00 ug/L	0.00
11) Chlorobenzene-d5	14.57	117	502552	25.00 ug/L	0.00
12) 1,4-Dichlorobenzene-d4	17.60	152	270883	25.00 ug/L	0.00

Target Compounds					Qvalue
2) Acetonitrile	6.69	41	110761	409.7980 ug/L	97
3) 3-Chloro-1-propene	7.07	41	3406131	382.6353 ug/L	82
4) 2-Chloro-1,3-butadiene	8.45	53	4356691	387.0705 ug/L	85
5) Ethyl Acetate	9.09	43	1116818	416.6642 ug/L	94
6) Methacrylonitrile	9.25	67	615482	418.2517 ug/L #	62
7) Isobutyl Alcohol	9.26	43	75472	842.9862 ug/L	97
9) Methyl methacrylate	11.42	41	1199317	414.7153 ug/L	71
10) 2-Nitropropane	11.75	43	538580	430.7760 ug/L	94
13) Cyclohexanone	15.84	55	45874	414.0575 ug/L #	55

(#) = qualifier out of range (m) = manual integration
8M412327.D A9FOOWT.M Thu May 19 09:01:31 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412327.D Vial: 9
Acq On : 17 May 2016 13:10 Operator: TMB
Sample : WG569079-08 400ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 9:00 2016 Quant Results File: A9FOOWT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration



8M412327.D A9FOOWT.M Thu May 19 09:01:31 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412328.D Vial: 10
Acq On : 17 May 2016 13:39 Operator: TMB
Sample : WG569079-09 500ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 09:01:31 2016 Quant Results File: A9FOOWT.RES

Quant Method : K:\ORGANICS\V...\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	10.71	96	695694	25.00	ug/L	0.00
11) Chlorobenzene-d5	14.57	117	495257	25.00	ug/L	0.00
12) 1,4-Dichlorobenzene-d4	17.60	152	266442	25.00	ug/L	0.00

Target Compounds						Qvalue
2) Acetonitrile	6.69	41	132245	492.6753	ug/L	99
3) 3-Chloro-1-propene	7.06	41	4228273	478.2834	ug/L	82
4) 2-Chloro-1,3-butadiene	8.46	53	5390856	482.2693	ug/L	85
5) Ethyl Acetate	9.09	43	1321922	496.6017	ug/L	95
6) Methacrylonitrile	9.25	67	729852	499.4083	ug/L #	61
7) Isobutyl Alcohol	9.26	43	85200	958.2367	ug/L	96
9) Methyl methacrylate	11.42	41	1416438	493.1877	ug/L	72
10) 2-Nitropropane	11.75	43	637335	513.2956	ug/L	93
13) Cyclohexanone	15.84	55	52070	477.8160	ug/L #	61

(#) = qualifier out of range (m) = manual integration
8M412328.D A9FOOWT.M Thu May 19 09:01:32 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412328.D Vial: 10
Acq On : 17 May 2016 13:39 Operator: TMB
Sample : WG569079-09 500ug/L A9/FOO STD 8260 Inst : HPMS8
Misc : 1,1 STD76072 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 9:00 2016 Quant Results File: A9FOOWT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration



8M412328.D A9FOOWT.M Thu May 19 09:01:32 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412331.D Vial: 13
Acq On : 17 May 2016 15:06 Operator: TMB
Sample : WG569079-10 100ug/L ALT SRC STD 8260 Inst : HPMS8
Misc : 1,1 STD75801 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 09:01:32 2016 Quant Results File: A9FOOWT.RES

Quant Method : K:\ORGANICS\V...\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	10.71	96	676765	25.00	ug/L	0.00
11) Chlorobenzene-d5	14.57	117	486447	25.00	ug/L	0.00
12) 1,4-Dichlorobenzene-d4	17.60	152	263036	25.00	ug/L	0.00

Target Compounds						Qvalue
2) Acetonitrile	6.69	41	29876	114.4154	ug/L	97
3) 3-Chloro-1-propene	7.07	41	858035	99.7718	ug/L	83
4) 2-Chloro-1,3-butadiene	8.45	53	1224194	112.5803	ug/L	87
5) Ethyl Acetate	9.09	43	295545	114.1317	ug/L	96
6) Methacrylonitrile	9.25	67	161574	113.6509	ug/L #	63
7) Isobutyl Alcohol	9.26	43	19948	230.6285	ug/L	94
9) Methyl methacrylate	11.42	41	326465	116.8508	ug/L	72
10) 2-Nitropropane	11.75	43	131261	108.6716	ug/L	94
13) Cyclohexanone	15.83	55	10459	97.2189	ug/L #	74

(#) = qualifier out of range (m) = manual integration
8M412331.D A9FOOWT.M Thu May 19 09:01:33 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412331.D Vial: 13
Acq On : 17 May 2016 15:06 Operator: TMB
Sample : WG569079-10 100ug/L ALT SRC STD 8260 Inst : HPMS8
Misc : 1,1 STD75801 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 19 9:00 2016 Quant Results File: A9FOOWT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\051716\8M412331.D Vial: 13
Acq On : 17 May 2016 15:06 Operator: TMB
Sample : WG569079-10 100ug/L ALT SRC STD 8260 Inst : HPMS8
Misc : 1,1 STD75801 Multiplr: 1.00
MS Integration Params: rteint.p

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 08:54:31 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 1% Max. R.T. Dev 0.50min
Max. RRF Dev : 75% Max. Rel. Area : 200%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	95	0.00
2	T	Acetonitrile	100.0000	114.4154	-14.4	109	0.00
3	T	3-Chloro-1-propene	100.0000	99.7718	0.2	91	0.00
4	T	2-Chloro-1,3-butadiene	100.0000	112.5803	-12.6	103	0.00
5	T	Ethyl Acetate	100.0000	114.1317	-14.1	104	0.00
6	T	Methacrylonitrile	100.0000	113.6509	-13.7	104	0.00
7	T	Isobutyl Alcohol	200.0000	230.6285	-15.3	111	0.00
8	T	1-Butanol	-1.0000	0.0000	0.0	0	0.01
9	T	Methyl methacrylate	100.0000	116.8508	-16.9	109	0.00
10	T	2-Nitropropane	100.0000	108.6716	-8.7	108	0.00
11	I	Chlorobenzene-d5	25.0000	25.0000	0.0	96	0.00
12	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	96	0.00
13	T	Cyclohexanone	100.0000	97.2189	2.8	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412331.D A9FOOWT.M Thu May 19 09:03:22 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413032.D Vial: 2
Acq On : 16 Jun 2016 14:25 Operator: FJB
Sample : WG573037-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 20 09:11:28 2016 Quant Results File: 8260WT.RES

Quant Method : K:\ORGANICS\V...\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	733857	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	538326	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	307491	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	215897	27.3819 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 109.52%		
43) 1,2-Dichloroethane-d4	10.30	65	234596	27.7486 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 111.00%		
58) Toluene-d8	12.68	98	742553	26.3564 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 105.44%		
80) p-Bromofluorobenzene	16.08	95	315298	24.7636 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 99.04%		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	709296	57.6485 ug/L	100
3) Chloromethane	3.62	50	334378	46.5578 ug/L	99
4) Vinyl Chloride	3.84	62	469712	55.1239 ug/L	99
5) 1,3-Butadiene	3.89	54	449731	62.9364 ug/L	97
6) Bromomethane	4.73	94	187928	36.9767 ug/L	99
7) Chloroethane	4.89	64	229428	48.9325 ug/L	97
8) Trichlorofluoromethane	5.37	101	947245	59.7435 ug/L	99
9) Diethyl ether	5.91	59	359174	94.8900 ug/L	97
10) Isoprene	5.94	67	490773	50.9718 ug/L	97
11) Acrolein	6.14	56	15092	44.5317 ug/L	97
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	413268	55.4511 ug/L	96
13) Acetone	6.24	43	43062	52.1548 ug/L	100
14) 1,1-Dichloroethene	6.47	61	648526	54.6514 ug/L	98
15) Tert-Butyl Alcohol	6.57	59	62702	236.9793 ug/L	98
16) Dimethyl Sulfide	6.73	62	205530	48.4405 ug/L	98
17) Iodomethane	6.99	142	88941	28.4929 ug/L	92
18) Methyl acetate	7.00	43	95456	42.7704 ug/L	98
19) Methylene Chloride	7.26	84	368680	48.8209 ug/L	100
20) Carbon Disulfide	7.29	76	1316779	53.0253 ug/L	100
21) Acrylonitrile	7.43	53	50980	47.5881 ug/L	94
22) Methyl Tert Butyl Ether	7.46	73	815898	49.9201 ug/L	99
23) trans-1,2-Dichloroethene	7.70	61	567448	52.0503 ug/L	100
24) n-Hexane	7.78	57	440578	54.4236 ug/L	99
25) Diisopropyl ether	8.12	45	1629728	95.9026 ug/L	96
26) Vinyl Acetate	8.30	43	383450	49.2832 ug/L	99
27) 1,1-Dichloroethane	8.33	63	708177	52.3061 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	1837543	100.9239 ug/L	98
29) 2-Butanone	8.88	43	55723	45.9348 ug/L	98
30) Propionitrile	8.99	54	34053	96.3269 ug/L	98
31) 2,2-Dichloropropane	9.11	77	881786	61.8026 ug/L	100
32) cis-1,2-Dichloroethene	9.17	96	429176	50.9485 ug/L	97
33) Chloroform	9.37	83	876461	54.8264 ug/L	100
34) 1-Bromopropane	9.51	122	69827	57.8711 ug/L	99
35) Bromochloromethane	9.61	130	221472	51.2707 ug/L	96
36) Tetrahydrofuran	9.63	42	65588	82.0898 ug/L	92
38) 1,1,1-Trichloroethane	9.91	97	904390	59.5586 ug/L	99
39) Cyclohexane	9.94	56	543065	51.6569 ug/L	96
40) 1,1-Dichloropropene	10.11	75	631550	54.5654 ug/L	100
41) Tert-Amyl-Methyl ether	10.21	73	1667474	100.8347 ug/L	100
42) Carbon Tetrachloride	10.25	117	809209	60.0125 ug/L	100

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413032.D Vial: 2
Acq On : 16 Jun 2016 14:25 Operator: FJB
Sample : WG573037-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 20 09:11:28 2016 Quant Results File: 8260WT.RES

Quant Method : K:\ORGANICS\V...\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc	Unit		Qvalue
45)	1,2-Dichloroethane	10.43	62	583022	56.6636	ug/L		98
46)	Benzene	10.46	78	1567451	51.9972	ug/L		99
47)	Trichloroethene	11.21	130	433837	53.9600	ug/L		99
48)	Methylcyclohexane	11.30	83	676953	55.4235	ug/L		100
49)	1,2-Dichloropropane	11.43	63	326457	48.6480	ug/L		86
50)	Bromodichloromethane	11.73	83	653209	55.3826	ug/L		100
51)	1,4-Dioxane	11.72	88	7428	220.3548	ug/L		98
52)	Dibromomethane	11.81	93	197458	49.7082	ug/L		99
53)	2-Chloroethyl Vinyl Ether	12.03	63	129844	45.0178	ug/L		95
54)	4-Methyl-2-Pentanone	12.06	58	53646	45.5327	ug/L		94
55)	cis-1,3-Dichloropropene	12.36	75	639780	53.2922	ug/L		100
56)	Dimethyl Disulfide	12.63	79	328801	51.3144	ug/L		99
59)	Toluene	12.78	91	1681041	50.3002	ug/L		100
60)	Ethyl Methacrylate	12.88	69	300733	46.8562	ug/L		98
62)	trans-1,3-Dichloropropene	12.96	75	572847	51.1483	ug/L		100
63)	1,1,2-Trichloroethane	13.18	97	244197	46.3057	ug/L		99
64)	2-Hexanone	13.12	58	49168	44.3674	ug/L	#	100
65)	1,3-Dichloropropane	13.48	76	441493	47.0219	ug/L		97
66)	Tetrachloroethene	13.60	164	383463	53.5544	ug/L		98
67)	Dibromochloromethane	13.87	129	388793	50.4917	ug/L		99
68)	1,2-Dibromoethane	14.13	107	247883	47.1634	ug/L		98
69)	1-Chlorohexane	14.21	91	594465	52.1092	ug/L		98
70)	Chlorobenzene	14.62	112	1121470	49.7630	ug/L		100
71)	1,1,1,2-Tetrachloroethane	14.66	131	456454	52.2807	ug/L		100
72)	Ethylbenzene	14.66	106	622088	50.4524	ug/L		99
73)	m-,p-Xylene	14.75	106	1491287	103.1131	ug/L		100
74)	o-Xylene	15.31	106	724388	51.2307	ug/L		99
75)	Styrene	15.35	104	1206191	52.0104	ug/L		96
76)	Bromoform	15.84	173	232298	50.8617	ug/L		99
77)	Isopropylbenzene	15.73	105	2026613	53.9527	ug/L		100
79)	1,1,2,2-Tetrachloroethane	15.95	83	256105	41.4928	ug/L		99
81)	1,2,3-Trichloropropane	16.14	110	85494	44.9807	ug/L		72
82)	trans-1,4-Dichloro-2-Buten	16.18	53	84374	43.6303	ug/L	#	51
83)	n-Propylbenzene	16.24	91	2386990	49.7580	ug/L		99
84)	Bromobenzene	16.37	156	473221	47.9586	ug/L		95
85)	1,3,5-Trimethylbenzene	16.43	105	1753288	50.1733	ug/L		100
86)	2-Chlorotoluene	16.52	91	1599256	47.7491	ug/L		100
87)	4-Chlorotoluene	16.56	91	1503375	49.6923	ug/L		100
88)	a-Methylstyrene	16.84	118	786649	47.4481	ug/L		98
89)	tert-Butylbenzene	16.90	134	334560	48.9377	ug/L		99
90)	1,2,4-Trimethylbenzene	16.95	105	1801178	50.1269	ug/L		100
91)	sec-Butylbenzene	17.17	105	2069862	50.3494	ug/L		99
92)	p-Isopropyltoluene	17.32	119	1766300	50.6022	ug/L		99
93)	1,3-Dichlorobenzene	17.52	146	942205	49.5900	ug/L		98
94)	1,4-Dichlorobenzene	17.64	146	911590	47.5896	ug/L		99
95)	n-Butylbenzene	17.85	91	1742674	50.5007	ug/L		99
96)	1,2-Dichlorobenzene	18.14	146	787813	46.6205	ug/L		99
97)	1,2-Dibromo-3-Chloropropan	19.13	75	54489	43.9791	ug/L		98
98)	1,2,4-Trichlorobenzene	20.28	180	613274	48.5529	ug/L		100
99)	Hexachlorobutadiene	20.43	225	322898	54.7126	ug/L		99
100)	Naphthalene	20.65	128	816338	43.5006	ug/L		100
101)	1,2,3-Trichlorobenzene	20.96	180	483344	45.3576	ug/L		99

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413032.D Vial: 2
Acq On : 16 Jun 2016 14:25 Operator: FJB
Sample : WG573037-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 20 9:11 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413032.D Vial: 2
Acq On : 16 Jun 2016 14:25 Operator: FJB
Sample : WG573037-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	1.0000	1.0000	0.0	94	0.00
2	T	Dichlorodifluoromethane	0.4192	0.4833	-15.3	104	0.00
3	P	Chloromethane	0.2447	0.2278	6.9	90	0.00
4	C	Vinyl Chloride	0.2903	0.3200	-10.2	103	0.00
5	T	1,3-Butadiene	0.2434	0.3064	-25.9#	115	0.00
6	T	Bromomethane	0.1654	0.1280	22.6	72	0.00
7	T	Chloroethane	0.1597	0.1563	2.1	94	0.00
8	T	Trichlorofluoromethane	0.5401	0.6454	-19.5	113	0.00
9	T	Diethyl ether	0.1290	0.1224	5.1	92	0.00
10	T	Isoprene	0.3280	0.3344	-1.9	99	0.00
11	T	Acrolein	0.0116	0.0103	11.0	86	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	0.2539	0.2816	-10.9	107	0.00
13	T	Acetone	0.0281	0.0293	-4.3	94	0.00
14	C	1,1-Dichloroethene	0.4043	0.4419	-9.3	105	0.00
15	T	Tert-Butyl Alcohol	0.0090	0.0107	-18.5	113	0.00
16	T	Dimethyl Sulfide	0.1445	0.1400	3.1	93	0.00
17	T	Iodomethane	0.0866	0.0606	30.0#	56	0.00
18	T	Methyl acetate	0.0760	0.0650	14.5	83	0.00
19	T	Methylene Chloride	0.2573	0.2512	2.4	95	0.00
20	T	Carbon Disulfide	0.8460	0.8972	-6.0	104	0.00
21	T	Acrylonitrile	0.0365	0.0347	4.8	85	0.00
22	T	Methyl Tert Butyl Ether	0.5568	0.5559	0.2	90	0.00
23	T	trans-1,2-Dichloroethene	0.3714	0.3866	-4.1	99	0.00
24	T	n-Hexane	0.2758	0.3002	-8.8	105	0.00
25	T	Diisopropyl ether	0.5789	0.5552	4.1	92	0.00
26	T	Vinyl Acetate	0.2651	0.2613	1.4	89	0.00
27	P	1,1-Dichloroethane	0.4612	0.4825	-4.6	98	0.00
28	T	Ethyl-Tert-Butyl ether	0.6203	0.6260	-0.9	94	0.00
29	T	2-Butanone	0.0413	0.0380	8.1	84	0.00
30	T	Propionitrile	0.0120	0.0116	3.7	89	0.00
31	T	2,2-Dichloropropane	0.4860	0.6008	-23.6	115	0.00
32	T	cis-1,2-Dichloroethene	0.2870	0.2924	-1.9	98	0.00
33	C	Chloroform	0.5446	0.5972	-9.7	104	0.00
34		1-Bromopropane	0.0411	0.0476	-15.8	110	0.00
35	T	Bromochloromethane	0.1472	0.1509	-2.5	96	0.00
36	T	Tetrahydrofuran	0.0272	0.0223	17.9	83	0.00
37	S	Dibromofluoromethane	0.2686	0.2942	-9.5	102	0.00
38	T	1,1,1-Trichloroethane	0.5173	0.6162	-19.1	111	0.00
39	T	Cyclohexane	0.3581	0.3700	-3.3	101	0.00
40	T	1,1-Dichloropropene	0.3943	0.4303	-9.1	103	0.00
41	T	Tert-Amyl-Methyl ether	0.5634	0.5681	-0.8	94	0.00
42	T	Carbon Tetrachloride	0.4593	0.5513	-20.0	112	0.00
43	S	1,2-Dichloroethane-d4	0.2880	0.3197	-11.0	102	0.00
44		Heptane	0.0000	0.0000	0.0	0#	-2.54#
45	T	1,2-Dichloroethane	0.3505	0.3972	-13.3	104	0.00
46	T	Benzene	1.0269	1.0680	-4.0	98	0.00
47	T	Trichloroethene	0.2739	0.2956	-7.9	102	0.00
48	T	Methylcyclohexane	0.4161	0.4612	-10.8	107	0.00
49	C	1,2-Dichloropropane	0.2286	0.2224	2.7	93	0.00
50	T	Bromodichloromethane	0.4018	0.4451	-10.8	100	0.00
51	T	1,4-Dioxane	0.0012	0.0013	-10.4	98	0.00
52	T	Dibromomethane	0.1353	0.1345	0.6	94	0.00
53	T	2-Chloroethyl Vinyl Ether	0.0983	0.0885	10.0	80	0.00
54	T	4-Methyl-2-Pentanone	0.0401	0.0365	8.9	83	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413032.D Vial: 2
Acq On : 16 Jun 2016 14:25 Operator: FJB
Sample : WG573037-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	0.4090	0.4359	-6.6	96	0.00
56	T	Dimethyl Disulfide	0.2183	0.2240	-2.6	96	0.00
57	I	Chlorobenzene-d5	1.0000	1.0000	0.0	100	0.00
58	S	Toluene-d8	1.3084	1.3794	-5.4	103	0.00
59	C	Toluene	1.5520	1.5614	-0.6	100	0.00
60	T	Ethyl Methacrylate	0.2981	0.2793	6.3	89	0.00
61		Paraldehyde	0.0000	0.0000	0.0	0#	-13.14#
62	T	trans-1,3-Dichloropropene	0.5201	0.5321	-2.3	96	0.00
63	T	1,1,2-Trichloroethane	0.2449	0.2268	7.4	91	0.00
64	T	2-Hexanone	0.0515	0.0457	11.3	86	0.00
65	T	1,3-Dichloropropane	0.4360	0.4101	6.0	90	0.00
66	T	Tetrachloroethene	0.3325	0.3562	-7.1	109	0.00
67	T	Dibromochloromethane	0.3576	0.3611	-1.0	97	0.00
68	T	1,2-Dibromoethane	0.2441	0.2302	5.7	91	0.00
69	T	1-Chlorohexane	0.5298	0.5521	-4.2	106	0.00
70	P	Chlorobenzene	1.0466	1.0416	0.5	100	0.00
71	T	1,1,1,2-Tetrachloroethane	0.4055	0.4240	-4.6	102	0.00
72	C	Ethylbenzene	0.5726	0.5778	-0.9	103	0.00
73	T	m-,p-Xylene	0.6716	0.6926	-3.1	103	0.00
74	T	o-Xylene	0.6566	0.6728	-2.5	102	0.00
75	T	Styrene	1.0770	1.1203	-4.0	100	0.00
76	P	Bromoform	0.2121	0.2158	-1.7	97	0.00
77	T	Isopropylbenzene	1.7444	1.8823	-7.9	106	0.00
78	I	1,4-Dichlorobenzene-d4	1.0000	1.0000	0.0	107	0.00
79	P	1,1,2,2-Tetrachloroethane	0.5018	0.4164	17.0	87	0.00
80	S	p-Bromofluorobenzene	1.0352	1.0254	0.9	108	0.00
81	T	1,2,3-Trichloropropane	0.1545	0.1390	10.0	96	0.00
82	T	trans-1,4-Dichloro-2-Butene	0.1376	0.1372	0.3	97	0.00
83	T	n-Propylbenzene	3.9003	3.8814	0.5	105	0.00
84	T	Bromobenzene	0.8022	0.7695	4.1	103	0.00
85	T	1,3,5-Trimethylbenzene	2.8411	2.8510	-0.3	106	0.00
86	T	2-Chlorotoluene	2.7231	2.6005	4.5	99	0.00
87	T	4-Chlorotoluene	2.4597	2.4446	0.6	112	0.00
88	T	a-Methylstyrene	1.3479	1.2791	5.1	102	0.00
89	T	tert-Butylbenzene	0.5558	0.5440	2.1	109	0.00
90	T	1,2,4-Trimethylbenzene	2.9214	2.9288	-0.3	107	0.00
91	T	sec-Butylbenzene	3.3424	3.3657	-0.7	108	0.00
92	T	p-Isopropyltoluene	2.8379	2.8721	-1.2	109	0.00
93	T	1,3-Dichlorobenzene	1.5448	1.5321	0.8	107	0.00
94	T	1,4-Dichlorobenzene	1.5574	1.4823	4.8	105	0.00
95	T	n-Butylbenzene	2.8056	2.8337	-1.0	108	0.00
96	T	1,2-Dichlorobenzene	1.3739	1.2810	6.8	103	0.00
97	T	1,2-Dibromo-3-Chloropropane	0.1007	0.0886	12.0	91	0.00
98	T	1,2,4-Trichlorobenzene	1.0269	0.9972	2.9	105	0.00
99	T	Hexachlorobutadiene	0.4798	0.5251	-9.4	122	0.00
100	T	Naphthalene	1.5257	1.3274	13.0	91	0.00
101	T	1,2,3-Trichlorobenzene	0.8664	0.7860	9.3	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M413032.D 8260WT.M Mon Jun 20 09:11:49 2016 Page 2

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413032.D Vial: 2
Acq On : 16 Jun 2016 14:25 Operator: FJB
Sample : WG573037-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	94	0.00
2	T	Dichlorodifluoromethane	50.0000	57.6485	-15.3	104	0.00
3	P	Chloromethane	50.0000	46.5578	6.9	90	0.00
4	C	Vinyl Chloride	50.0000	55.1240	-10.2	103	0.00
5	T	1,3-Butadiene	50.0000	62.9364	-25.9#	115	0.00
6	T	Bromomethane	50.0000	36.9767	26.0#	72	0.00
7	T	Chloroethane	50.0000	48.9325	2.1	94	0.00
8	T	Trichlorofluoromethane	50.0000	59.7435	-19.5	113	0.00
9	T	Diethyl ether	100.0000	94.8900	5.1	92	0.00
10	T	Isoprene	50.0000	50.9718	-1.9	99	0.00
11	T	Acrolein	50.0000	44.5317	10.9	86	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	50.0000	55.4511	-10.9	107	0.00
13	T	Acetone	50.0000	52.1548	-4.3	94	0.00
14	C	1,1-Dichloroethene	50.0000	54.6514	-9.3	105	0.00
15	T	Tert-Butyl Alcohol	200.0000	236.9793	-18.5	113	0.00
16	T	Dimethyl Sulfide	50.0000	48.4405	3.1	93	0.00
17	T	Iodomethane	50.0000	28.4930	43.0#	56	0.00
18	T	Methyl acetate	50.0000	42.7704	14.5	83	0.00
19	T	Methylene Chloride	50.0000	48.8209	2.4	95	0.00
20	T	Carbon Disulfide	50.0000	53.0253	-6.1	104	0.00
21	T	Acrylonitrile	50.0000	47.5881	4.8	85	0.00
22	T	Methyl Tert Butyl Ether	50.0000	49.9201	0.2	90	0.00
23	T	trans-1,2-Dichloroethene	50.0000	52.0503	-4.1	99	0.00
24	T	n-Hexane	50.0000	54.4236	-8.8	105	0.00
25	T	Diisopropyl ether	100.0000	95.9026	4.1	92	0.00
26	T	Vinyl Acetate	50.0000	49.2832	1.4	89	0.00
27	P	1,1-Dichloroethane	50.0000	52.3061	-4.6	98	0.00
28	T	Ethyl-Tert-Butyl ether	100.0000	100.9239	-0.9	94	0.00
29	T	2-Butanone	50.0000	45.9348	8.1	84	0.00
30	T	Propionitrile	100.0000	96.3269	3.7	89	0.00
31	T	2,2-Dichloropropane	50.0000	61.8026	-23.6	115	0.00
32	T	cis-1,2-Dichloroethene	50.0000	50.9485	-1.9	98	0.00
33	C	Chloroform	50.0000	54.8264	-9.7	104	0.00
34		1-Bromopropane	50.0000	57.8711	-15.7	110	0.00
35	T	Bromochloromethane	50.0000	51.2707	-2.5	96	0.00
36	T	Tetrahydrofuran	100.0000	82.0898	17.9	83	0.00
37	S	Dibromofluoromethane	25.0000	27.3819	-9.5	102	0.00
38	T	1,1,1-Trichloroethane	50.0000	59.5586	-19.1	111	0.00
39	T	Cyclohexane	50.0000	51.6569	-3.3	101	0.00
40	T	1,1-Dichloropropene	50.0000	54.5654	-9.1	103	0.00
41	T	Tert-Amyl-Methyl ether	100.0000	100.8347	-0.8	94	0.00
42	T	Carbon Tetrachloride	50.0000	60.0125	-20.0	112	0.00
43	S	1,2-Dichloroethane-d4	25.0000	27.7486	-11.0	102	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	50.0000	56.6636	-13.3	104	0.00
46	T	Benzene	50.0000	51.9972	-4.0	98	0.00
47	T	Trichloroethene	50.0000	53.9600	-7.9	102	0.00
48	T	Methylcyclohexane	50.0000	55.4235	-10.8	107	0.00
49	C	1,2-Dichloropropane	50.0000	48.6480	2.7	93	0.00
50	T	Bromodichloromethane	50.0000	55.3826	-10.8	100	0.00
51	T	1,4-Dioxane	200.0000	220.3548	-10.2	98	0.00
52	T	Dibromomethane	50.0000	49.7082	0.6	94	0.00
53	T	2-Chloroethyl Vinyl Ether	50.0000	45.0178	10.0	80	0.00
54	T	4-Methyl-2-Pentanone	50.0000	45.5327	8.9	83	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413032.D Vial: 2
Acq On : 16 Jun 2016 14:25 Operator: FJB
Sample : WG573037-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	50.0000	53.2922	-6.6	96	0.00
56	T	Dimethyl Disulfide	50.0000	51.3144	-2.6	96	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	25.0000	26.3564	-5.4	103	0.00
59	C	Toluene	50.0000	50.3002	-0.6	100	0.00
60	T	Ethyl Methacrylate	50.0000	46.8562	6.3	89	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	50.0000	51.1483	-2.3	96	0.00
63	T	1,1,2-Trichloroethane	50.0000	46.3057	7.4	91	0.00
64	T	2-Hexanone	50.0000	44.3674	11.3	86	0.00
65	T	1,3-Dichloropropane	50.0000	47.0219	6.0	90	0.00
66	T	Tetrachloroethene	50.0000	53.5544	-7.1	109	0.00
67	T	Dibromochloromethane	50.0000	50.4917	-1.0	97	0.00
68	T	1,2-Dibromoethane	50.0000	47.1634	5.7	91	0.00
69	T	1-Chlorohexane	50.0000	52.1093	-4.2	106	0.00
70	P	Chlorobenzene	50.0000	49.7630	0.5	100	0.00
71	T	1,1,1,2-Tetrachloroethane	50.0000	52.2807	-4.6	102	0.00
72	C	Ethylbenzene	50.0000	50.4524	-0.9	103	0.00
73	T	m-,p-Xylene	100.0000	103.1131	-3.1	103	0.00
74	T	o-Xylene	50.0000	51.2307	-2.5	102	0.00
75	T	Styrene	50.0000	52.0104	-4.0	100	0.00
76	P	Bromoform	50.0000	50.8617	-1.7	97	0.00
77	T	Isopropylbenzene	50.0000	53.9527	-7.9	106	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	107	0.00
79	P	1,1,2,2-Tetrachloroethane	50.0000	41.4928	17.0	87	0.00
80	S	p-Bromofluorobenzene	25.0000	24.7636	0.9	108	0.00
81	T	1,2,3-Trichloropropane	50.0000	44.9807	10.0	96	0.00
82	T	trans-1,4-Dichloro-2-Butene	50.0000	43.6303	12.7	97	0.00
83	T	n-Propylbenzene	50.0000	49.7580	0.5	105	0.00
84	T	Bromobenzene	50.0000	47.9586	4.1	103	0.00
85	T	1,3,5-Trimethylbenzene	50.0000	50.1733	-0.3	106	0.00
86	T	2-Chlorotoluene	50.0000	47.7491	4.5	99	0.00
87	T	4-Chlorotoluene	50.0000	49.6923	0.6	112	0.00
88	T	a-Methylstyrene	50.0000	47.4481	5.1	102	0.00
89	T	tert-Butylbenzene	50.0000	48.9378	2.1	109	0.00
90	T	1,2,4-Trimethylbenzene	50.0000	50.1269	-0.3	107	0.00
91	T	sec-Butylbenzene	50.0000	50.3494	-0.7	108	0.00
92	T	p-Isopropyltoluene	50.0000	50.6022	-1.2	109	0.00
93	T	1,3-Dichlorobenzene	50.0000	49.5900	0.8	107	0.00
94	T	1,4-Dichlorobenzene	50.0000	47.5896	4.8	105	0.00
95	T	n-Butylbenzene	50.0000	50.5007	-1.0	108	0.00
96	T	1,2-Dichlorobenzene	50.0000	46.6205	6.8	103	0.00
97	T	1,2-Dibromo-3-Chloropropane	50.0000	43.9791	12.0	91	0.00
98	T	1,2,4-Trichlorobenzene	50.0000	48.5529	2.9	105	0.00
99	T	Hexachlorobutadiene	50.0000	54.7126	-9.4	122	0.00
100	T	Naphthalene	50.0000	43.5006	13.0	91	0.00
101	T	1,2,3-Trichlorobenzene	50.0000	45.3576	9.3	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M413032.D 8260WT.M Mon Jun 20 09:11:52 2016 Page 2

2.1.1.5 Raw QC Data

BFB

```
Data File : C:\MSDCHEM\2\DATA\051316\8M412218.D          Vial: 3
Acq On    : 13 May 2016  12:36                       Operator: TMB
Sample    : WG568561-01 50ng BFB STD 8260            Inst    : HPMS8
Misc      : 1,1 STD76034                             Multiplr: 1.00
MS Integration Params: rteint.p
Method    : C:\MSDCHEM\2\METHODS\BFB.M (RTE Integrator)
Title     : SOP:OVL MSV01
```





AutoFind: Scans 102, 103, 104; Background Corrected with Scan 97

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	6575	PASS
75	95	30	60	51.3	20171	PASS
95	95	100	100	100.0	39312	PASS
96	95	5	9	6.7	2649	PASS
173	174	0.00	2	0.4	122	PASS
174	95	50	100	85.5	33616	PASS
175	174	5	9	8.2	2746	PASS
176	174	95	101	99.2	33360	PASS
177	176	5	9	6.8	2267	PASS

8M412218.D BFB.M Fri May 13 12:47:24 2016

BFB

```
Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413031.D Vial: 1
Acq On    : 16 Jun 2016  14:02                        Operator: FJB
Sample    : WG573037-01 BFB 50ng 8260                 Inst    : HPMS8
Misc      : 1,1 STD76496                              Multiplr: 1.00
MS Integration Params: RTEINT.P
Method    : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title     : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
```





AutoFind: Scans 101, 102, 103; Background Corrected with Scan 97

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	6352	PASS
75	95	30	60	57.2	18249	PASS
95	95	100	100	100.0	31890	PASS
96	95	5	9	7.4	2366	PASS
173	174	0.00	2	1.5	352	PASS
174	95	50	100	73.7	23499	PASS
175	174	5	9	8.0	1870	PASS
176	174	95	101	97.3	22869	PASS
177	176	5	9	6.7	1538	PASS

8M413031.D 8260WT.M Mon Jun 20 09:08:01 2016

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413034.D Vial: 4
Acq On : 16 Jun 2016 15:23 Operator: FJB
Sample : WG573038-01 BLANK 6/16 8260 Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 08:18:21 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	717429	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	533754	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	292043	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	209977	27.2409 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 108.96%		
43) 1,2-Dichloroethane-d4	10.30	65	235474	28.4902 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 113.96%		
58) Toluene-d8	12.68	98	740524	26.5096 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 106.04%		
80) p-Bromofluorobenzene	16.08	95	307463	25.4256 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 101.72%		
Target Compounds					Qvalue
36) Tetrahydrofuran	9.64	42	956	1.2239 ug/L	79

(#) = qualifier out of range (m) = manual integration
8M413034.D 8260WT.M Mon Jun 20 09:19:27 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413034.D Vial: 4
Acq On : 16 Jun 2016 15:23 Operator: FJB
Sample : WG573038-01 BLANK 6/16 8260 Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 8:18 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413034.D 8260WT.M Mon Jun 20 09:19:28 2016 Page 2



#36
Tetrahydrofuran
Concen: 1.22 ug/L
RT: 9.64 min Scan# 712
Delta R.T. 0.01 min
Lab File: 8M413034.D
Acq: 16 Jun 2016 15:23

Tgt Ion: 42 Resp: 956

Ion	Ratio	Lower	Upper
42	100		
72	74.6	33.8	78.8
71	41.9	32.0	74.8



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413035.D Vial: 5
Acq On : 16 Jun 2016 15:52 Operator: FJB
Sample : WG573038-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76597 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 08:18:21 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	724153	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	540053	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	307888	25.00 ug/L	0.00

System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	213643	27.4591 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 109.84%		
43) 1,2-Dichloroethane-d4	10.31	65	245621	29.4420 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 117.76%		
58) Toluene-d8	12.68	98	736706	26.0653 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 104.28%		
80) p-Bromofluorobenzene	16.08	95	316242	24.8057 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 99.24%		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	252315	20.7819 ug/L	99
3) Chloromethane	3.62	50	134208	18.9371 ug/L	99
4) Vinyl Chloride	3.85	62	192999	22.9533 ug/L	100
5) 1,3-Butadiene	3.88	54	21596	3.0627 ug/L	93
6) Bromomethane	4.72	94	67601	14.0319 ug/L	96
7) Chloroethane	4.89	64	94578	20.4419 ug/L	99
8) Trichlorofluoromethane	5.38	101	348634	22.2833 ug/L	99
9) Diethyl ether	5.90	59	376158	100.7087 ug/L	98
10) Isoprene	5.93	67	202345	21.2972 ug/L	96
11) Acrolein	6.13	56	30585	91.4560 ug/L	92
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	157632	21.4340 ug/L	96
13) Acetone	6.23	43	17743	21.7775 ug/L	94
14) 1,1-Dichloroethene	6.46	61	221517	18.9174 ug/L	99
15) Tert-Butyl Alcohol	6.58	59	54296	207.9591 ug/L	97
16) Dimethyl Sulfide	6.73	62	149005	35.5890 ug/L	99
17) Iodomethane	6.99	142	33488	12.6554 ug/L	96
18) Methyl acetate	7.00	43	45239	20.5416 ug/L	98
19) Methylene Chloride	7.26	84	143716	19.2860 ug/L	97
20) Carbon Disulfide	7.29	76	443914	18.1155 ug/L	99
21) Acrylonitrile	7.44	53	20303	19.2061 ug/L	97
22) Methyl Tert Butyl Ether	7.46	73	362621	22.4840 ug/L	97
23) trans-1,2-Dichloroethene	7.70	61	216143	20.0918 ug/L	99
24) n-Hexane	7.78	57	148742	18.6200 ug/L	98
25) Diisopropyl ether	8.12	45	1674940	99.8839 ug/L	96
26) Vinyl Acetate	8.30	43	113665	14.8046 ug/L	99
27) 1,1-Dichloroethane	8.32	63	260446	19.4944 ug/L	98
28) Ethyl-Tert-Butyl ether	8.70	59	1819224	101.2567 ug/L	98
29) 2-Butanone	8.88	43	24177	20.1972 ug/L	96
30) Propionitrile	8.99	54	32822	94.0889 ug/L	98
31) 2,2-Dichloropropane	9.11	77	323361	22.9674 ug/L	100
32) cis-1,2-Dichloroethene	9.17	96	170309	20.4887 ug/L	97
33) Chloroform	9.37	83	334264	21.1899 ug/L	100
34) 1-Bromopropane	9.51	122	36217	30.4181 ug/L	100
35) Bromochloromethane	9.60	130	88133	20.6762 ug/L	92
36) Tetrahydrofuran	9.62	42	66083	83.8177 ug/L	91
38) 1,1,1-Trichloroethane	9.91	97	336501	22.4572 ug/L	100
39) Cyclohexane	9.94	56	203672	19.6331 ug/L	95
40) 1,1-Dichloropropene	10.10	75	235033	20.5788 ug/L	98
41) Tert-Amyl-Methyl ether	10.21	73	1758569	107.7684 ug/L	100
42) Carbon Tetrachloride	10.25	117	299046	22.4750 ug/L	98

(#) = qualifier out of range (m) = manual integration
8M413035.D 8260WT.M Mon Jun 20 09:08:47 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413035.D Vial: 5
Acq On : 16 Jun 2016 15:52 Operator: FJB
Sample : WG573038-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76597 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 08:18:21 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc	Unit		Qvalue
45)	1,2-Dichloroethane	10.42	62	236926	23.3353	ug/L		98
46)	Benzene	10.46	78	605690	20.3618	ug/L		99
47)	Trichloroethene	11.21	130	172682	21.7657	ug/L		99
48)	Methylcyclohexane	11.30	83	251284	20.8488	ug/L		98
49)	1,2-Dichloropropane	11.42	63	134844	20.3635	ug/L		90
50)	Bromodichloromethane	11.73	83	251163	21.5803	ug/L		98
51)	1,4-Dioxane	11.72	88	5908	177.6120	ug/L		98
52)	Dibromomethane	11.81	93	77825	19.8542	ug/L		99
53)	2-Chloroethyl Vinyl Ether	12.03	63	55535	19.5124	ug/L		91
54)	4-Methyl-2-Pentanone	12.05	58	22978	19.7642	ug/L		93
55)	cis-1,3-Dichloropropene	12.36	75	273539	23.0905	ug/L		100
56)	Dimethyl Disulfide	12.62	79	137758	21.7873	ug/L		95
59)	Toluene	12.79	91	661287	19.7238	ug/L		99
60)	Ethyl Methacrylate	12.88	69	133238	20.6930	ug/L		98
62)	trans-1,3-Dichloropropene	12.96	75	234780	20.8960	ug/L		100
63)	1,1,2-Trichloroethane	13.18	97	103597	19.5817	ug/L		97
64)	2-Hexanone	13.12	58	20672	18.5940	ug/L	#	95
65)	1,3-Dichloropropane	13.48	76	194941	20.6961	ug/L		95
66)	Tetrachloroethene	13.60	164	144676	20.1408	ug/L		99
67)	Dibromochloromethane	13.87	129	158132	20.4706	ug/L		100
68)	1,2-Dibromoethane	14.12	107	105951	20.0943	ug/L		99
69)	1-Chlorohexane	14.21	91	223061	19.4904	ug/L		97
70)	Chlorobenzene	14.63	112	460997	20.3904	ug/L		99
71)	1,1,1,2-Tetrachloroethane	14.66	131	182348	20.8187	ug/L		99
72)	Ethylbenzene	14.66	106	245573	19.8527	ug/L		100
73)	m-,p-Xylene	14.75	106	592898	40.8641	ug/L		95
74)	o-Xylene	15.31	106	299422	21.1082	ug/L		99
75)	Styrene	15.35	104	480813	20.6661	ug/L		96
76)	Bromoform	15.83	173	91438	19.9564	ug/L		99
77)	Isopropylbenzene	15.73	105	813144	21.5784	ug/L		100
79)	1,1,2,2-Tetrachloroethane	15.95	83	109337	17.6914	ug/L		99
81)	1,2,3-Trichloropropane	16.14	110	37733	19.8268	ug/L		73
82)	trans-1,4-Dichloro-2-Buten	16.19	53	28275	15.0111	ug/L		99
83)	n-Propylbenzene	16.24	91	981466	20.4328	ug/L		100
84)	Bromobenzene	16.37	156	189420	19.1720	ug/L		95
85)	1,3,5-Trimethylbenzene	16.42	105	710541	20.3071	ug/L		100
86)	2-Chlorotoluene	16.51	91	681730	20.3282	ug/L		98
87)	4-Chlorotoluene	16.56	91	596710	19.6981	ug/L		99
88)	a-Methylstyrene	16.83	118	344648	20.7612	ug/L		100
89)	tert-Butylbenzene	16.90	134	134805	19.6932	ug/L		98
90)	1,2,4-Trimethylbenzene	16.94	105	716254	19.9077	ug/L		99
91)	sec-Butylbenzene	17.16	105	833548	20.2499	ug/L		99
92)	p-Isopropyltoluene	17.32	119	700643	20.0466	ug/L		100
93)	1,3-Dichlorobenzene	17.51	146	379985	19.9735	ug/L		100
94)	1,4-Dichlorobenzene	17.64	146	378307	19.7240	ug/L		99
95)	n-Butylbenzene	17.85	91	671796	19.4428	ug/L		100
96)	1,2-Dichlorobenzene	18.14	146	333637	19.7182	ug/L		97
97)	1,2-Dibromo-3-Chloropropan	19.13	75	24498	19.7473	ug/L		99
98)	1,2,4-Trichlorobenzene	20.28	180	244999	19.3716	ug/L		99
99)	Hexachlorobutadiene	20.43	225	125683	21.2686	ug/L		98
100)	Naphthalene	20.65	128	326897	17.3970	ug/L		99
101)	1,2,3-Trichlorobenzene	20.96	180	202292	18.9589	ug/L		97

(#) = qualifier out of range (m) = manual integration
8M413035.D 8260WT.M Mon Jun 20 09:08:47 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413035.D Vial: 5
Acq On : 16 Jun 2016 15:52 Operator: FJB
Sample : WG573038-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76597 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 8:18 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413035.D 8260WT.M Mon Jun 20 09:08:48 2016 Page 3

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413036.D Vial: 6
Acq On : 16 Jun 2016 16:21 Operator: FJB
Sample : WG573038-03 20ug/L LCS2 8260 Inst : HPMS8
Misc : 1,1 STD76597 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 08:18:22 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	735322	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	553122	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	312159	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	213768	27.0578 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 108.24%		
43) 1,2-Dichloroethane-d4	10.30	65	240558	28.3971 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 113.60%		
58) Toluene-d8	12.68	98	752327	25.9891 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 103.96%		
80) p-Bromofluorobenzene	16.08	95	319552	24.7224 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 98.88%		
Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	268839	21.8065 ug/L	99
3) Chloromethane	3.62	50	145372	20.2008 ug/L	99
4) Vinyl Chloride	3.85	62	205956	24.1222 ug/L	100
5) 1,3-Butadiene	3.88	54	22591	3.1551 ug/L	94
6) Bromomethane	4.72	94	76970	15.6580 ug/L	99
7) Chloroethane	4.89	64	102039	21.7196 ug/L	98
8) Trichlorofluoromethane	5.38	101	362939	22.8452 ug/L	98
9) Diethyl ether	5.90	59	397097	104.6998 ug/L	97
10) Isoprene	5.94	67	218692	22.6682 ug/L	99
11) Acrolein	6.14	56	33320	98.1209 ug/L	96
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	161414	21.6149 ug/L	94
13) Acetone	6.23	43	19963	24.1301 ug/L	97
14) 1,1-Dichloroethene	6.46	61	233526	19.6401 ug/L	99
15) Tert-Butyl Alcohol	6.57	59	65029	245.2844 ug/L	97
16) Dimethyl Sulfide	6.73	62	155773	36.6403 ug/L	99
17) Iodomethane	6.99	142	48971	16.9853 ug/L	95
18) Methyl acetate	7.00	43	45727	20.4478 ug/L	97
19) Methylene Chloride	7.25	84	150567	19.8985 ug/L	100
20) Carbon Disulfide	7.29	76	462391	18.5829 ug/L	100
21) Acrylonitrile	7.43	53	21986	20.4823 ug/L	92
22) Methyl Tert Butyl Ether	7.47	73	371382	22.6774 ug/L	98
23) trans-1,2-Dichloroethene	7.70	61	222944	20.4092 ug/L	100
24) n-Hexane	7.78	57	159355	19.6456 ug/L	98
25) Diisopropyl ether	8.12	45	1757628	103.2229 ug/L	96
26) Vinyl Acetate	8.30	43	118636	15.2174 ug/L	100
27) 1,1-Dichloroethane	8.32	63	269975	19.9007 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	1896750	103.9682 ug/L	98
29) 2-Butanone	8.88	43	25340	20.8472 ug/L	94
30) Propionitrile	8.99	54	36109	101.9393 ug/L	100
31) 2,2-Dichloropropane	9.10	77	336668	23.5494 ug/L	99
32) cis-1,2-Dichloroethene	9.17	96	179241	21.2357 ug/L	98
33) Chloroform	9.37	83	347237	21.6779 ug/L	100
34) 1-Bromopropane	9.51	122	37842	31.3001 ug/L	100
35) Bromochloromethane	9.61	130	89039	20.5714 ug/L	96
36) Tetrahydrofuran	9.63	42	72974	91.1522 ug/L	97
38) 1,1,1-Trichloroethane	9.91	97	353698	23.2464 ug/L	100
39) Cyclohexane	9.93	56	214088	20.3237 ug/L	97
40) 1,1-Dichloropropene	10.11	75	243549	21.0005 ug/L	99
41) Tert-Amyl-Methyl ether	10.21	73	1837419	110.8902 ug/L	99
42) Carbon Tetrachloride	10.25	117	311990	23.0917 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M413036.D 8260WT.M Mon Jun 20 09:08:50 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413036.D Vial: 6
Acq On : 16 Jun 2016 16:21 Operator: FJB
Sample : WG573038-03 20ug/L LCS2 8260 Inst : HPMS8
Misc : 1,1 STD76597 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 08:18:22 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
45)	1,2-Dichloroethane	10.42	62	242552	23.5265 ug/L		98
46)	Benzene	10.46	78	636475	21.0718 ug/L		100
47)	Trichloroethene	11.21	130	181127	22.4834 ug/L		99
48)	Methylcyclohexane	11.30	83	268536	21.9418 ug/L		99
49)	1,2-Dichloropropane	11.42	63	141319	21.0171 ug/L		89
50)	Bromodichloromethane	11.73	83	265829	22.4935 ug/L		99
51)	1,4-Dioxane	11.72	88	6780	200.7309 ug/L		99
52)	Dibromomethane	11.81	93	79309	19.9255 ug/L		98
53)	2-Chloroethyl Vinyl Ether	12.02	63	56761	19.6402 ug/L		94
54)	4-Methyl-2-Pentanone	12.06	58	24424	20.6889 ug/L		92
55)	cis-1,3-Dichloropropene	12.36	75	282091	23.4507 ug/L		99
56)	Dimethyl Disulfide	12.62	79	140775	21.9263 ug/L		94
59)	Toluene	12.79	91	690537	20.1096 ug/L		99
60)	Ethyl Methacrylate	12.88	69	134927	20.4602 ug/L		98
62)	trans-1,3-Dichloropropene	12.96	75	242695	21.0901 ug/L		99
63)	1,1,2-Trichloroethane	13.18	97	106261	19.6107 ug/L		98
64)	2-Hexanone	13.12	58	20293	17.8218 ug/L	#	97
65)	1,3-Dichloropropane	13.48	76	202255	20.9653 ug/L		97
66)	Tetrachloroethene	13.60	164	150723	20.4869 ug/L		99
67)	Dibromochloromethane	13.87	129	164885	20.8405 ug/L		100
68)	1,2-Dibromoethane	14.12	107	107229	19.8562 ug/L		100
69)	1-Chlorohexane	14.21	91	238806	20.3731 ug/L		97
70)	Chlorobenzene	14.63	112	479877	20.7240 ug/L		100
71)	1,1,1,2-Tetrachloroethane	14.66	131	187840	20.9390 ug/L		99
72)	Ethylbenzene	14.66	106	252056	19.8954 ug/L		96
73)	m-,p-Xylene	14.75	106	620374	41.7475 ug/L		96
74)	o-Xylene	15.31	106	310957	21.4035 ug/L		98
75)	Styrene	15.35	104	503519	21.1307 ug/L		97
76)	Bromoform	15.83	173	93776	19.9830 ug/L		98
77)	Isopropylbenzene	15.73	105	848704	21.9899 ug/L		99
79)	1,1,2,2-Tetrachloroethane	15.95	83	113419	18.1008 ug/L		100
81)	1,2,3-Trichloropropane	16.14	110	37740	19.5591 ug/L		73
82)	trans-1,4-Dichloro-2-Buten	16.19	53	34462	17.9212 ug/L		82
83)	n-Propylbenzene	16.24	91	1024017	21.0269 ug/L		99
84)	Bromobenzene	16.37	156	197176	19.6840 ug/L		97
85)	1,3,5-Trimethylbenzene	16.42	105	742680	20.9352 ug/L		99
86)	2-Chlorotoluene	16.51	91	701689	20.6371 ug/L		98
87)	4-Chlorotoluene	16.56	91	615763	20.0490 ug/L		99
88)	a-Methylstyrene	16.83	118	365485	21.7152 ug/L		99
89)	tert-Butylbenzene	16.89	134	141823	20.4349 ug/L		98
90)	1,2,4-Trimethylbenzene	16.94	105	742233	20.3475 ug/L		100
91)	sec-Butylbenzene	17.16	105	865092	20.7287 ug/L		100
92)	p-Isopropyltoluene	17.32	119	739719	20.8751 ug/L		99
93)	1,3-Dichlorobenzene	17.51	146	398135	20.6412 ug/L		100
94)	1,4-Dichlorobenzene	17.64	146	395783	20.3529 ug/L		99
95)	n-Butylbenzene	17.85	91	695375	19.8498 ug/L		99
96)	1,2-Dichlorobenzene	18.14	146	349070	20.3481 ug/L		99
97)	1,2-Dibromo-3-Chloropropan	19.13	75	26890	21.3789 ug/L		96
98)	1,2,4-Trichlorobenzene	20.28	180	257503	20.0816 ug/L		99
99)	Hexachlorobutadiene	20.43	225	129798	21.6644 ug/L		98
100)	Naphthalene	20.65	128	344850	18.1014 ug/L		100
101)	1,2,3-Trichlorobenzene	20.96	180	209338	19.3508 ug/L		98

(#) = qualifier out of range (m) = manual integration
8M413036.D 8260WT.M Mon Jun 20 09:08:51 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\061616\8M413036.D Vial: 6
Acq On : 16 Jun 2016 16:21 Operator: FJB
Sample : WG573038-03 20ug/L LCS2 8260 Inst : HPMS8
Misc : 1,1 STD76597 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 17 8:18 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



2.2 Semivolatiles Data

2.2.1 Semivolatiles GC/MS Data (Dioxane)

2.2.1.1 Summary Data

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-01	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	126-061416	**Prep Method:**	3520C	**Prep Date:**	06/15/2016 18:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573365	**Analyst:**	MES	**Run Date:**	06/21/2016 18:40
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	15M18480
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	73.6	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	C08-061416	**Prep Method:**	3520C	**Prep Date:**	06/15/2016 18:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573365	**Analyst:**	MES	**Run Date:**	06/21/2016 19:02
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	15M18481
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	44.3	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	C03-061416	**Prep Method:**	3520C	**Prep Date:**	06/15/2016 18:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573365	**Analyst:**	MES	**Run Date:**	06/21/2016 19:25
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	15M18482
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	43.2	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.2.1.2 QC Summary Data

Example 8270 Calculations

1.0 Calculating the Response Factor (RF) from the initial calibration (ICAL) data:

RF = [(Ax) (Cis)] / [(Ais) (Cx)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured:	1261197
Cis = Concentration of the specific internal standard (ug/mL)	40
Ais = Area of the characteristic ion of the specific internal standard	608044
Cx = Concentration of the compound in the standard being measured (ug/mL)	50
RF = Calculated Response Factor	**1.65935**

2.0 Calculating the concentration (C) of a compound in water using the data from the prep log and quantitation report: *

Cx = [(Ax) (Cis) (Vf) (D)] / [(Ais) (RF) (Vi)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured	367250
Cis = Concentration of the specific internal standard (ug/mL)	40
Vf = Final volume of sample extract from prep log (mL)	1
D = Dilution factor for sample as a multiplier (10x = 10)	1
Ais = Area of the characteristic ion of the specific internal standard	511641
RF = Average RF from the ICAL	1.65935
Vi = Initial volume of sample extracted from prep log (mL)	1021
Cx = Concentration of the compound in the sample being measured (ug/mL)	0.016947
Cx = Concentration of the compound in the sample being measured (ug/L)	**16.947**

3.0 Calculating the concentration (C) of a compound in soil using the data from the prep log and quantitation report: *

Cx = [(Ax) (Cis) (Vf) (D)] / [(Ais) (RF) (Wi)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured	367250
Cis = Concentration of the specific internal standard (ug/mL)	40
Vf = Final volume of sample extract from prep log (mL)	1
D = Dilution factor for sample as a multiplier (10x = 10)	1
Ais = Area of the characteristic ion of the specific internal standard	511641
RF = Average RF from the ICAL	1.65935
Wi = Initial weight of sample extracted (g) from prep log	30
Cx = Concentration of the compound in the sample being measured (ug/g)	0.576763
Cx = Concentration of the compound in the sample being measured (ug/kg)	**576.7627**

Dry weight correction:

Percent solids (PCT_S)	50
Cd = (Cx) (100)/PCT_S	**1153.525** ug/kg

* Concentrations appearing on the instrument quantitation reports are on-column results and do not take into account initial volume, final volume, and the dilution factor.

4.0 Concentration from Linear Regression

Step 1: Retrieve Curve Data From Plot, y = mx + b

y = response ratio = response of analyte / response of IS = Ax/Ais

x = amount ratio = concentration analyte/concentration internal standard = Cx / Cis

m = slope from curve plot

b = intercept from curve plot

Step 2: Calculate y from Quantitation Report

y = 16790/784838 = 0.02139

Step 3: Solve for x

x = (y - b)/m = [(0.02139 -(- 0.0435)]/ 0.0783 = 0.829

Step 4: Solve for analyte concentration Cx

Cx = Cis (x) = (25.0)(0.829) = 20.72 ug/L

Example Spreadsheet Calculation:

Slope from curve, m:	0.0783	
Intercept from curve, b:	-0.0435	
Area of analyte, Ax:	16790	
Area of Internal Standard , Ais:	784484	
Concentration of IS, Cis	25.00	ug/L
Response Ratio (y) :	0.021403	
Amount Ratio:	0.828897	
Concentration (Cx):	20.72241	ug/L

5.0 Concentration from Quadratic Regression

Step 1 - Retrieve Curve Data from Plot, y = Ax^2 + Bx + C
Where:
Ax^2 + Bx + (C - y) = 0
A, B, C = constants from the ICAL quadratic regression
y = Response ratio = Area of analyte/Area of internal standard (IS)
x = Amount ratio = Concentration of analyte/concentration of IS

Step 2: Calculate y from Quantitation Report

y = Ax/Ais

Step 3: Solve for x using the quadratic formula
Ax^2 + Bx + C - y = 0

$$x = \frac{b \pm \sqrt{\left(b^2 - 4a(c - y)\right)}}{2a}$$ (Two possible solutions)

Step 4: Solve for analyte concentration Cx

Cx = (Cis)(Amount ratio)

Example Spreadsheet Calculation:

Value of A from plot:	**0.0259**	
Value of B from plot:	**0.0596**	
Value of C from plot:	**-0.0165**	
Area of analyte from quantitation report:	**203233**	
Area of IS from quantiation report:	**1425653**	
Response ratio, y:	**0.142554**	
C - y:	**-0.15905**	
Root 1 - Computed amount ratio , X1:	**-3.88278**	
Root 2 - Computed amount ratio , X2:	**1.581623**	use this solution
Concentration of IS, Cis:	**40.00**	
Concentration of analyte, Cx:	**63.26**	ug/L

Microbac Laboratories Inc.
Continuous Sample Extract Log

Workgroup: WG572806
Analyst: CPD
Spike Analyst: CPD
Method: 3520C
Run Date: 06/15/2016 18:01
SOP: EXB01 Revison 20
Spike Witness: JDH
Surr Solution: STD74051

TIME ON: 19:15 OFF: 13:30 ON: 17:50 OFF: 12:00

CH2CL2 Lot #: COA18982
sodium sulfate Lot #: COA18814
10N NaOH Lot #: RGT37076
1:1 H2SO4 Lot #: RGT36755

	SAMPLE #	Type	Reference	Prod	pH	Init Amnt	Surr Amnt	Spike Amnt	Spike Sol	Final Vol	Color
1	L16060730-01	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Colored
2	L16060730-02	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Colored
3	L16060730-03	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
4	L16060756-04	SAMP		827-DIOX	<2>12	1000 mL	.05 mL			1 mL	Colored
5	L16060809-03	SAMP		827-DIOX	<2>12	1000 mL	.05 mL			1 mL	Colored
6	L16060822-01	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Colored
7	L16060822-03	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
8	L16060822-04	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
9	L16060822-05	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
10	L16060823-01	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
11	L16060823-03	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Colored
12	L16060823-04	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
13	L16060854-01	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Colored
14	WG572806-01	BLANK		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
15	WG572806-02	LCS		827-DIOXANE	<2>12	1000 mL	.05 mL	.05 mL	STD75256	1 mL	Transparent
16	WG572806-03	LCS2		827-DIOXANE	<2>12	1000 mL	.05 mL	.05 mL	STD75256	1 mL	Transparent

Due to insufficient sample volume, this preparation batch failed to include the method prescribed MS and MSD.

LCS TV3 Cell 6
LCS DUP TV3 Cell 1

Analyst: Chelsie Davis

Reviewer: Chris Hill

EXTRACT - Modified 01/23/2009
PDF ID: 4813908
Report generated: 06/20/2016 08:50

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062116
Analyst1: MES Analyst2: NA
Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:
Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG573365

Internal STD: STD76627 Surrogate STD: NA Calibration STD

CCV STD: STD74784 LCS STD: STD74214 MS/MSD STD:

Comments:

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	15M18455	BAKEOUT	1	1		06/21/16 08:46
2	15M18456	WG573501-01 5PPM DFTPP STD	1	1	STD75016	06/21/16 09:05
3	15M18457	WG573501-01 5PPM DFTPP STD	1	1	STD75016	06/21/16 09:34
4	15M18458	WG573501-02 5PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 09:52
5	15M18459	WG573501-03 10 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 10:27
6	15M18460	WG573501-04 7.5 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 10:49
7	15M18461	WG573501-05 2.5 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 11:12
8	15M18462	WG573501-06 1 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 11:34
9	15M18463	WG573501-07 0.4 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 11:57
10	15M18464	WG573501-08 5PPM 1,4-DIOXANE ALT SRC	1	1	STD74214	06/21/16 12:20
11	15M18465	WG572806-01 BLANK 6/15	1	1		06/21/16 12:54
12	15M18466	WG572806-02 LCS 6/15	1	1		06/21/16 13:17
13	15M18467	WG572806-03 LCSDUP 6/15	1	1		06/21/16 13:40
14	15M18468	WG573008-01 BLANK 6/17	7	1	SOIL	06/21/16 14:03
15	15M18469	WG573008-02 LCS 6/17	7	1	SOIL	06/21/16 14:26
16	15M18470	WG573008-03 LCSDUP 6/17	7	1	SOIL	06/21/16 14:48
17	15M18471	L16060730-01	1	1		06/21/16 15:11
18	15M18472	L16060730-02	1	1		06/21/16 15:34
19	15M18473	L16060854-01	1	1		06/21/16 15:57
20	15M18474	L16060730-03	1	1		06/21/16 16:19
21	15M18475	L16060756-04 827-DIOX	1	1		06/21/16 16:46
22	15M18476	L16060822-01	1	1		06/21/16 17:08
23	15M18477	L16060822-03	1	1		06/21/16 17:31
24	15M18478	L16060822-04	1	1		06/21/16 17:54
25	15M18479	L16060822-05	1	1		06/21/16 18:17
26	15M18480	L16060823-01	1	1		06/21/16 18:40
27	15M18481	L16060823-03	1	1		06/21/16 19:02
28	15M18482	L16060823-04	1	1		06/21/16 19:25
29	15M18483	L16060809-03 827-DIOX	1	1		06/21/16 19:48
30	15M18484	L16060854-01 2X	1	2		06/21/16 20:10
31	15M18485	BAKEOUT	1	1		06/21/16 20:33
32	15M18486	BAKEOUT	1	1		06/21/16 20:55
33	15M18487	BAKEOUT	1	1		06/21/16 21:18

Page: 1 Approved: 22-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062116

Analyst1: MES Analyst2: NA

Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:

Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG573365

Internal STD: STD76627 Surrogate STD: NA

CCV STD: STD74784 LCS STD: STD74214

Comments

Seq.	Rerun	Dil.	Reason	Analytes
2				
	WG573501-01 5PPM DFTPP STD rerun DFTPP, pentachlorophenol tailing >2.			
14				
	WG573008-01 BLANK 6/17 DNR-not needed.			
15				
	WG573008-02 LCS DNR-not needed.			
16				
	WG573008-03 LCSDUP DNR-not needed.			
19	X	2	Over Calibration Range	
	L16060854-01			

Page: 2 Approved: 22-JUN-16

Microbac Laboratories Inc.

Data Checklist

Date: 21-JUN-2016
Analyst: MES
Analyst: NA
Method: 827-DIOX
Instrument: HPMS15
Curve Workgroup: NA
Runlog ID: 75850
Analytical Workgroups: L16060730, 060756, 060809, 060822, 060823, 060854

ANALYTICAL	
System Performance Check	X
DFTPP (MS)	X
Endrin/DDT breakdown (8081/MS)	X
Pentachlorophenol/benzidine tailing (MS)	X
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	X
Average RF	X
Linear regression or higher order curve	NA
Alternate source standard (ICV) % Difference	X
Continuing Calibration (CCV)	NA
% D/% Drift	NA
Minimum response factors (MS)	NA
Continuing calibration blank (CCB) (IC)	NA
Special standards	NA
Blanks	X
TCL hits	X
Surrogate recoveries	X
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	X
MS/MSD/Sample duplicates	NA
Recoveries	NA
%RPD	NA
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	X
Surrogate recoveries	X
Internal standard areas (MS)	X
Library searches (MS)	NA
Calculations & correct factors	X
Compounds above calibration range	X
Reruns	X
Manual integrations	NA
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	NA
Check for completeness	X
Primary Reviewer	MES
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	ECL

Primary Reviewer: 22-JUN-2016

Secondary Reviewer: 22-JUN-2016

CHECKLIST1 - Modified 03/05/2008

Generated: JUN-22-2016 16:00:51

Microbac Laboratories Inc.

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 8270D

AAB#: WG573365

Login Number: L16060823

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
126-061416	01	06/14/16					06/15/2016	1.3	7		06/21/16	6	40	
C08-061416	03	06/14/16					06/15/2016	1.2	7		06/21/16	6	40	
C03-061416	04	06/14/16					06/15/2016	1.2	7		06/21/16	6.1	40	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4819086
Report generated 06/23/2016 13:28

Microbac Laboratories Inc.
SURROGATE STANDARDS

Login Number: L16060823
Instrument Id: HPMS15
Workgroup (AAB#): WG573365

Method: 827-DIOXANE
CAL ID: HPMS15-21-JUN-16
Matrix: Water

Sample Number	Dilution	Tag	1
L16060823-01	1.00	01	73.6
L16060823-03	1.00	01	44.3
L16060823-04	1.00	01	43.2
WG572806-01	1.00	02	61.7
WG572806-02	1.00	02	62.5
WG572806-03	1.00	02	46.5

Surrogates	Surrogate Limits		
1 - 1,4-Dioxane-d8	20	-	129

Underline = Result out of surrogate limits

DL = surrogate diluted out
ND = surrogate not detected

SURROGATES - Modified 03/06/2008
PDF File ID: 4819093
Report generated: 06/23/2016 13:28

METHOD BLANK SUMMARY

Login Number: L16060823
Blank File ID: 15M18465
Prep Date: 06/15/16 18:01
Analyzed Date: 06/21/16 12:54
Analyst: MES

Work Group: WG573365
Blank Sample ID: WG572806-01
Instrument ID: HPMS15
Method: 8270D

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG572806-02	15M18466	06/21/16 13:17	02
LCS2	WG572806-03	15M18467	06/21/16 13:40	02
126-061416	L16060823-01	15M18480	06/21/16 18:40	01
C08-061416	L16060823-03	15M18481	06/21/16 19:02	01
C03-061416	L16060823-04	15M18482	06/21/16 19:25	01

Report Name: BLANK_SUMMARY
PDF File ID: 4821148
Report generated 06/23/2016 13:28

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060823 Prep Date: 06/15/16 18:01 Sample ID: WG572806-01
Instrument ID: HPMS15 Run Date: 06/21/16 12:54 Prep Method: 3520C
File ID: 15M18465 Analyst: MES Method: 8270D
Workgroup (AAB#): WG573365 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS15-21-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
1,4-Dioxane	0.500	2.00	0.500	1	U

Surrogates	% Recovery	Surrogate Limits	Qualifier
1,4-Dioxane-d8	61.7	20 - 129	PASS

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name:BLANK
PDF ID: 4819087
23-JUN-2016 13:28

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060823 Analyst: MES Prep Method: 3520C
Instrument ID: HPMS15 Matrix: Water Method: 8270D
Workgroup (AAB#): WG573365 Units: ug/L
QC Key: DOD4 Lot #: STD75256

Sample ID: WG572806-02 LCS File ID: 15M18466 Run Date: 06/21/2016 13:17
Sample ID: WG572806-03 LCS2 File ID: 15M18467 Run Date: 06/21/2016 13:40

Analytes	LCS			LCS2			%RPD	%Rec Limits	RPD Lmt	Q
	Known	Found	% REC	Known	Found	% REC				
1,4-Dioxane	5.00	3.03	60.6	5.00	2.27	45.5	28.6	30 - 104	30	

Surogates	LCS	LCS2	Surrogate Limits	Qualifier
	% Recovery	% Recovery		
1,4-Dioxane-d8	62.5	46.5	20 - 129	PASS

* EXCEEDS %REC LIMIT

\# EXCEEDS RPD LIMIT

LCS_LCS2 - Modified 03/06/2008
PDF File ID: 4819088
Report generated: 06/23/2016 13:28

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

DFTPP

Login Number: L16060823
Instrument: HPMS15
Analyst: MES
Workgroup: WG573501

Tune ID: WG573501-01
Run Date: 06/21/2016
Run Time: 09:34
File ID: 15M18457
Cal ID: HPMS15 - 21-JUN-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51.0	198	30.0	60.0	38.8	34488	PASS
68.0	69.0	0	2.00	1.36	500	PASS
69.0	198	0	100	41.4	36833	PASS
70.0	69.0	0	2.00	0.467	172	PASS
127	198	40.0	60.0	51.9	46131	PASS
197	198	0	1.00	0	0	PASS
198	198	100	100	100	88880	PASS
199	198	5.00	9.00	6.66	5918	PASS
275	198	10.0	30.0	24.9	22090	PASS
365	198	1.00	100	2.65	2358	PASS
441	443	0.0100	100	75.7	9586	PASS
442	198	40.0	100	73.4	65261	PASS
443	442	17.0	23.0	19.4	12659	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573501-02	STD-CCV	01	06/21/2016 09:52	
WG573501-03	STD	01	06/21/2016 10:27	
WG573501-04	STD	01	06/21/2016 10:49	
WG573501-05	STD	01	06/21/2016 11:12	
WG573501-06	STD	01	06/21/2016 11:34	
WG573501-07	STD	01	06/21/2016 11:57	
WG573501-08	SSCV	01	06/21/2016 12:20	
WG572806-01	BLANK	02	06/21/2016 12:54	
WG572806-02	LCS	02	06/21/2016 13:17	
WG572806-03	LCS2	02	06/21/2016 13:40	
L16060823-01	126-061416	01	06/21/2016 18:40	
L16060823-03	C08-061416	01	06/21/2016 19:02	
L16060823-04	C03-061416	01	06/21/2016 19:25	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4819091
Report generated 06/23/2016 13:28

Microbac Laboratories Inc.

INITIAL CALIBRATION SUMMARY

Login Number: L16060823
Analytical Method: 8270D
ICAL Workgroup: WG573501

Instrument ID: HPMS15
Initial Calibration Date: 21-JUN-16 11:57
Column ID: F

Analyte		AVG RF	% RSD	LINEAR (R)	QUAD (R^2)
1,4-Dioxane		0.3060	2.56		

R = Correlation coefficient; 0.995 minimum
R^2 = Coefficient of determination; 0.99 minimum

INT_CAL - Modified 03/06/2008
PDF File ID: 4819089
Report generated 06/23/2016 13:28

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16060823

Analytical Method: 8270D

Instrument ID: HPMS15

Initial Calibration Date: 21-JUN-16 11:57

Column ID: F

Analyte	WG573501-02			WG573501-03			WG573501-04		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
1,4-Dioxane	5.00	60883.0000	0.3073	10.0	138820.000	0.3164	7.50	103182.000	0.3106

INT_CAL - Modified 03/06/2008
PDF File ID: 4819089
Report generated 06/23/2016 13:28

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16060823

Analytical Method: 8270D

Instrument ID: HPMS15

Initial Calibration Date: 21-JUN-16 11:57

Column ID: F

Analyte	WG573501-05			WG573501-06			WG573501-07		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
1,4-Dioxane	2.50	26333.0000	0.2969	1.00	13250.0000	0.2966	0.400	5650.00000	0.3081

INT_CAL - Modified 03/06/2008
PDF File ID: 4819089
Report generated 06/23/2016 13:28

Calibration Table Report
Method: DIOXANE_D8.M
Title: OVD MSS01 SIM 1,4-dioxane ICAL 062116
Last Calibration: Tue Jun 21 12:51:17 2016
Curve:WG573501
Calibration Files

Compound	10 15M18459.D	7.5 15M18460.D	5 15M18458.D	2.5 15M18461.D	1 15M18462.D	0.4 15M18463.D	Avg	%RSD
1,4-Dichlorobenzene-d4	ISTD							
1,4-Dioxane	0.316	0.311	0.307	0.297	0.297	0.308	0.306	2.558
1,4-Dioxane-d8	0.426	0.418	0.417	0.402	0.395		0.411	3.044
Nitrobenzene-d5	1.202	1.173	1.128	1.002	0.927	0.906	1.056	12.163
2-Fluorobiphenyl	2.735	2.736	2.691	2.621	2.563	2.651	2.666	2.548
p-Terphenyl-d14	2.793	2.774	2.622	2.460	2.367	2.405	2.570	7.266

Wed Jun 22 12:17:33 2016

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16060823 Run Date: 06/21/2016 Sample ID: WG573501-08
Instrument ID: HPMS15 Run Time: 12:20 Method: 8270D
File ID: 15M18464 Analyst: MES QC Key: DOD4
ICal Workgroup: WG573501 Cal ID: HPMS15 - 21-JUN-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
1,4-Dioxane		5000	5050	ug/L	0.309	0.900	20	

* Exceeds %D Limit

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4819090
Report generated 06/23/2016 13:28

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060823
Instrument ID: HPMS15
Workgroup (AAB#): WG573365

ICAL CCV Number: WG573501-02
CAL ID: HPMS15-21-JUN-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1
WG573501-02	NA	NA	39637
Upper Limit	NA	NA	79274
Lower Limit	NA	NA	19819
L16060823-01	1.00	01	32097
L16060823-03	1.00	01	24731
L16060823-04	1.00	01	24008
WG572806-01	1.00	02	38899
WG572806-02	1.00	02	27126
WG572806-03	1.00	02	34071

IS-1 - 1,4-Dichlorobenzene-d4

Underline = Response outside limits

INTERNAL_STD_ICAL - Modified 03/06/2008
PDF File ID: 4819092
Report generated 06/23/2016 13:28

Microbac Laboratories Inc.

INTERNAL STANDARD RETENTION TIME SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060823
Instrument ID: HPMS15
Workgroup (AAB#): WG573365

ICAL CCV Number: WG573501-02
CAL ID: HPMS15 - 21-JUN-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1
WG573501-02	NA	NA	6.93
Upper Limit	NA	NA	7.43
Lower Limit	NA	NA	6.43
L16060823-01	1.00	01	6.936
L16060823-03	1.00	01	6.936
L16060823-04	1.00	01	6.936
WG572806-01	1.00	02	6.936
WG572806-02	1.00	02	6.933
WG572806-03	1.00	02	6.933

IS-1 - 1,4-Dichlorobenzene-d4

Underline = Response outside limits

INTERNAL_STD_RT_ICAL - Modified 03/06/2008
PDF File ID: 4821149
Report generated: 06/23/2016 13:28

2.2.1.3 Sample Data

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18480.D
Acq On : 21 Jun 2016 18:40
Operator : MES
Sample : L16060823-01
Misc : 1,1
ALS Vial : 23 Sample Multiplier: 1

Quant Time: Jun 22 12:26:34 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.936	152	32097	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.834	96	48618	3.6811	ug/L	0.00
Spiked Amount 5.000			Recovery	=	73.60%	
4) Nitrobenzene-d5	7.436	82	6000	0.1770	ug/L	0.00
Spiked Amount 5.000			Recovery	=	3.60%	
5) 2-Fluorobiphenyl	8.926	172	1168	0.0136	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.20%	
6) p-Terphenyl-d14	12.203	244	1355	0.0164	ug/L	0.02
Spiked Amount 5.000			Recovery	=	0.40%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18480.D
Acq On : 21 Jun 2016 18:40
Operator : MES
Sample : L16060823-01
Misc : 1,1
ALS Vial : 23 Sample Multiplier: 1

Quant Time: Jun 22 12:26:34 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18481.D
Acq On : 21 Jun 2016 19:02
Operator : MES
Sample : L16060823-03
Misc : 1,1
ALS Vial : 24 Sample Multiplier: 1

Quant Time: Jun 22 08:03:40 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.936	152	24731	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.820	96	22524	2.2134	ug/L	0.00
Spiked Amount 5.000			Recovery	=	44.20%	
4) Nitrobenzene-d5	7.436	82	2394	0.0916	ug/L	0.00
Spiked Amount 5.000			Recovery	=	1.80%	
5) 2-Fluorobiphenyl	8.926	172	86	0.0013	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.00%	
6) p-Terphenyl-d14	12.203	244	909	0.0143	ug/L	0.02
Spiked Amount 5.000			Recovery	=	0.20%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18481.D
Acq On : 21 Jun 2016 19:02
Operator : MES
Sample : L16060823-03
Misc : 1,1
ALS Vial : 24 Sample Multiplier: 1

Quant Time: Jun 22 08:03:40 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18482.D
Acq On : 21 Jun 2016 19:25
Operator : MES
Sample : L16060823-04
Misc : 1,1
ALS Vial : 25 Sample Multiplier: 1

Quant Time: Jun 22 08:03:43 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.936	152	24008	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.820	96	21324	2.1585	ug/L	0.00
Spiked Amount 5.000			Recovery	=	43.20%	
4) Nitrobenzene-d5	7.436	82	1135	0.0448	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.80%	
5) 2-Fluorobiphenyl	8.926	172	139	0.0022	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.00%	
6) p-Terphenyl-d14	12.188	244	796	0.0129	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.20%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18482.D
Acq On : 21 Jun 2016 19:25
Operator : MES
Sample : L16060823-04
Misc : 1,1
ALS Vial : 25 Sample Multiplier: 1

Quant Time: Jun 22 08:03:43 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



2.2.1.4 Standards Data

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18458.D
Acq On : 21 Jun 2016 9:52
Operator : MES
Sample : WG573501-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 22 08:04:39 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	6.933	152	39637	1.0000 ug/mL	# 0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.827	96	82573	5.0627 ug/L	0.00
Spiked Amount 5.000			Recovery	= 101.20%	
4) Nitrobenzene-d5	7.431	82	223381	5.3353 ug/L	0.00
Spiked Amount 5.000			Recovery	= 106.80%	
5) 2-Fluorobiphenyl	8.922	172	533101	5.0444 ug/L	0.00
Spiked Amount 5.000			Recovery	= 100.80%	
6) p-Terphenyl-d14	12.184	244	519544	5.1003 ug/L	0.00
Spiked Amount 5.000			Recovery	= 102.00%	
Target Compounds					Qvalue
2) 1,4-Dioxane	3.862	58	60883	5.0202 ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18458.D
Acq On : 21 Jun 2016 9:52
Operator : MES
Sample : WG573501-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 22 08:04:39 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18459.D
Acq On : 21 Jun 2016 10:27
Operator : MES
Sample : WG573501-03 10 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 22 08:04:42 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	43874	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.834	96	186767	10.3452	ug/L	0.00
Spiked Amount 5.000			Recovery = 207.00%			
4) Nitrobenzene-d5	7.431	82	527549	11.3834	ug/L	0.00
Spiked Amount 5.000			Recovery = 227.60%			
5) 2-Fluorobiphenyl	8.926	172	1199799	10.2565	ug/L	0.00
Spiked Amount 5.000			Recovery = 205.20%			
6) p-Terphenyl-d14	12.240	244	1225328	10.8672	ug/L	0.06
Spiked Amount 5.000			Recovery = 217.40%			
Target Compounds						Qvalue
2) 1,4-Dioxane	3.869	58	138820	10.3412	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18459.D
Acq On : 21 Jun 2016 10:27
Operator : MES
Sample : WG573501-03 10 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 22 08:04:42 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18460.D
Acq On : 21 Jun 2016 10:49
Operator : MES
Sample : WG573501-04 7.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 22 08:04:45 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	6.933	152	44299	1.0000 ug/mL	# 0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.834	96	138833	7.6163 ug/L	0.00
Spiked Amount 5.000			Recovery	= 152.40%	
4) Nitrobenzene-d5	7.431	82	389810	8.3306 ug/L	0.00
Spiked Amount 5.000			Recovery	= 166.60%	
5) 2-Fluorobiphenyl	8.923	172	909050	7.6965 ug/L	0.00
Spiked Amount 5.000			Recovery	= 154.00%	
6) p-Terphenyl-d14	12.203	244	921380	8.0931 ug/L	0.02
Spiked Amount 5.000			Recovery	= 161.80%	
Target Compounds					Qvalue
2) 1,4-Dioxane	3.876	58	103182	7.6127 ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18460.D
Acq On : 21 Jun 2016 10:49
Operator : MES
Sample : WG573501-04 7.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 22 08:04:45 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18461.D
Acq On : 21 Jun 2016 11:12
Operator : MES
Sample : WG573501-05 2.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jun 22 08:04:48 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	35478	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.827	96	35630	2.4406	ug/L	0.00
Spiked Amount 5.000			Recovery = 48.80%			
4) Nitrobenzene-d5	7.431	82	88869	2.3714	ug/L	0.00
Spiked Amount 5.000			Recovery = 47.40%			
5) 2-Fluorobiphenyl	8.923	172	232500	2.4579	ug/L	0.00
Spiked Amount 5.000			Recovery = 49.20%			
6) p-Terphenyl-d14	12.199	244	218191	2.3930	ug/L	0.02
Spiked Amount 5.000			Recovery = 47.80%			
Target Compounds						Qvalue
2) 1,4-Dioxane	3.862	58	26333	2.4259	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18461.D
Acq On : 21 Jun 2016 11:12
Operator : MES
Sample : WG573501-05 2.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jun 22 08:04:48 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18462.D
Acq On : 21 Jun 2016 11:34
Operator : MES
Sample : WG573501-06 1 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jun 22 08:04:51 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	44672	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.841	96	17660	0.9607	ug/L	0.01
Spiked Amount 5.000			Recovery	=	19.20%	
4) Nitrobenzene-d5	7.431	82	41409	0.8776	ug/L	0.00
Spiked Amount 5.000			Recovery	=	17.60%	
5) 2-Fluorobiphenyl	8.922	172	114491	0.9612	ug/L	0.00
Spiked Amount 5.000			Recovery	=	19.20%	
6) p-Terphenyl-d14	12.203	244	105713	0.9208	ug/L	0.02
Spiked Amount 5.000			Recovery	=	18.40%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.876	58	13250	0.9694	ug/L	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18462.D
Acq On : 21 Jun 2016 11:34
Operator : MES
Sample : WG573501-06 1 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jun 22 08:04:51 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18463.D
Acq On : 21 Jun 2016 11:57
Operator : MES
Sample : WG573501-07 0.4 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jun 22 08:04:54 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	45840	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.841	96	7629	0.4045	ug/L	0.01
Spiked Amount 5.000			Recovery	=	8.00%	
4) Nitrobenzene-d5	7.431	82	16604	0.3429	ug/L	0.00
Spiked Amount 5.000			Recovery	=	6.80%	
5) 2-Fluorobiphenyl	8.923	172	48617	0.3978	ug/L	0.00
Spiked Amount 5.000			Recovery	=	8.00%	
6) p-Terphenyl-d14	12.210	244	44097	0.3743	ug/L	0.03
Spiked Amount 5.000			Recovery	=	7.40%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.876	58	5650	0.4028	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18463.D
Acq On : 21 Jun 2016 11:57
Operator : MES
Sample : WG573501-07 0.4 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jun 22 08:04:54 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18464.D
Acq On : 21 Jun 2016 12:20
Operator : MES
Sample : WG573501-08 5PPM 1,4-DIOXANE ALT SRC
Misc : 1,1 STD74214
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 22 08:04:57 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	41944	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.792	96	22	0.0013	ug/L	-0.04
Spiked Amount 5.000			Recovery	=	0.00%	
4) Nitrobenzene-d5	7.436	82	56	0.0013	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.00%	
5) 2-Fluorobiphenyl	8.926	172	23	0.0002	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.00%	
6) p-Terphenyl-d14	12.214	244	128	0.0012	ug/L	0.03
Spiked Amount 5.000			Recovery	=	0.00%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.869	58	64750	5.0454	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18464.D
Acq On : 21 Jun 2016 12:20
Operator : MES
Sample : WG573501-08 5PPM 1,4-DIOXANE ALT SRC
Misc : 1,1 STD74214
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 22 08:04:57 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18464.D
Acq On : 21 Jun 2016 12:20
Operator : MES
Sample : WG573501-08 5PPM 1,4-DIOXANE ALT SRC
Misc : 1,1 STD74214
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 22 08:04:57 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	5.000	5.045	-0.9	106	0.00
3 S	1,4-Dioxane-d8	5.000	0.001	100.0#	0	-0.04
4 S	Nitrobenzene-d5	5.000	0.001	100.0#	0	0.00
5 S	2-Fluorobiphenyl	5.000	0.000	100.0#	0	0.00
6 S	p-Terphenyl-d14	5.000	0.001	100.0#	0	0.03

(#) = Out of Range SPCC's out = 0 CCC's out = 0

2.2.1.5 Raw QC Data

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\msdchem\1\methods\DFTPPQ.M
Title : DFTPP
Last Update : Mon Apr 27 13:16:06 2015



AutoFind: Scans 234, 235, 236; Background Corrected with Scan 228

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	38.8	34488	PASS
68	69	0.00	2	1.4	500	PASS
69	198	0.00	100	41.4	36833	PASS
70	69	0.00	2	0.5	172	PASS
127	198	40	60	51.9	46131	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	88880	PASS
199	198	5	9	6.7	5918	PASS
275	198	10	30	24.9	22090	PASS
365	198	1	100	2.7	2358	PASS
441	443	0.01	100	75.7	9586	PASS
442	198	40	100	73.4	65261	PASS
443	442	17	23	19.4	12659	PASS

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 21 09:47:12 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18457.D\data.ms

(1) Pentachlorophenol

7.054min (+ 0.105) 0.00 ug/ml

response 30738

Ion	Exp%	Act%
265.80	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 21 09:47:12 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18457.D\data.ms

(2) Benzidine

8.502min (+ 0.008) 0.00 ug/ml

response 532770

Ion	Exp%	Act%
184.10	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 21 09:47:12 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18457.D\data.ms

(3) DDT

9.486min (-0.108) 0.00 ug/ml

response 180187

Ion	Exp%	Act%
235.00	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18465.D
Acq On : 21 Jun 2016 12:54
Operator : MES
Sample : WG572806-01 BLANK 6/15
Misc : 1,1
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Jun 21 16:33:15 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.936	152	38899	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.834	96	49398	3.0862	ug/L	0.00
Spiked Amount 5.000			Recovery = 61.80%			
4) Nitrobenzene-d5	7.431	82	1663307	40.4808	ug/L	0.00
Spiked Amount 5.000			Recovery = 809.60%			
5) 2-Fluorobiphenyl	8.926	172	3076633	29.6644	ug/L	0.00
Spiked Amount 5.000			Recovery = 593.20%			
6) p-Terphenyl-d14	12.258	244	4064361	40.6562	ug/L	0.07
Spiked Amount 5.000			Recovery = 813.20%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18465.D
Acq On : 21 Jun 2016 12:54
Operator : MES
Sample : WG572806-01 BLANK 6/15
Misc : 1,1
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Jun 21 16:33:15 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18466.D
Acq On : 21 Jun 2016 13:17
Operator : MES
Sample : WG572806-02 LCS 6/15
Misc : 1,1
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jun 22 08:05:00 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	27126	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.820	96	34886	3.1254	ug/L	0.00
Spiked Amount 5.000			Recovery	=	62.60%	
4) Nitrobenzene-d5	7.436	82	2047	0.0714	ug/L	0.00
Spiked Amount 5.000			Recovery	=	1.40%	
5) 2-Fluorobiphenyl	8.923	172	526	0.0073	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.20%	
6) p-Terphenyl-d14	12.229	244	1164	0.0167	ug/L	0.04
Spiked Amount 5.000			Recovery	=	0.40%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.855	58	25161	3.0316	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18466.D
Acq On : 21 Jun 2016 13:17
Operator : MES
Sample : WG572806-02 LCS 6/15
Misc : 1,1
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jun 22 08:05:00 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18467.D
Acq On : 21 Jun 2016 13:40
Operator : MES
Sample : WG572806-03 LCSDUP 6/15
Misc : 1,1
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Jun 22 08:05:03 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	34071	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.834	96	32606	2.3257	ug/L	0.00
Spiked Amount 5.000			Recovery	=	46.60%	
4) Nitrobenzene-d5	7.238	82	1167	0.0324	ug/L	-0.19
Spiked Amount 5.000			Recovery	=	0.60%	
5) 2-Fluorobiphenyl	8.923	172	1594	0.0175	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.40%	
6) p-Terphenyl-d14	12.207	244	949	0.0108	ug/L	0.02
Spiked Amount 5.000			Recovery	=	0.20%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.869	58	23691	2.2726	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18467.D
Acq On : 21 Jun 2016 13:40
Operator : MES
Sample : WG572806-03 LCSDUP 6/15
Misc : 1,1
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Jun 22 08:05:03 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



2.2 General Chromatography Data

2.2.2 6850 LC/MS Data

2.2.2.1 Summary Data

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-01	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	126-061416	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/24/2016 21:44
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	1LM.LM35468
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	0.140	J	0.400	0.200	0.100

J	Estimated value ; the analyte concentration was less than the LOQ.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	C08-061416	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/24/2016 22:03
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	1LM.LM35469
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	0.200	U	0.400	0.200	0.100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	C03-061416	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/24/2016 22:22
Collect Date:	06/14/2016 14:15	**Dilution:**	10	**File ID:**	1LM.LM35470
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	5.48		4.00	2.00	1.00

2.2.2.2 QC Summary Data

Example Calculation 6850 - Perchlorate

Concentration from Linear Regression

Step 1: Retrieve Curve Data From Plot, y = mx + b

y = response ratio = response of analyte / response of internal standard (IS) = Rx/Ristd

x = amount ratio = concentration analyte/concentration internal standard (IS) = Cx / Cistd

m = slope from curve (1.45)

b = intercept from curve (-0.00242)

y = 1.45x + -0.00242

Step 2: Substitute the value for y

where y = 12600/226000 = 0.055752

Step 3: Solve for x

x = (y - b)/m = 0.0040119

Step 4: Solve for analyte concentration Cx

Cx = (Cis)(x) = (5 ug/L)(0.040119) = 0.200594 ug/L

Example Calculation - Water:

Slope from curve, m:	1.45
Intercept from curve, b:	-0.00242
Response of analyte, Rx:	12600
Response of Internal Standard , Ristd:	226000
Concentration of IS, Cistd (ug/L):	5.00
Response Ratio:	0.05575
Amount Ratio:	0.04012
Analyte Concentration, Cx (ug/L) :	0.200594

Example Calculation - Soil:

Analyte Concentration, Cx (ug/L):	0.20059
Amount of soil extracted (g):	5.00
Final volume of extract (mL):	50.00
Percent solids (Pct wt.)	100
Concentration in soil (ug/kg):	2.005938

Perchlorate Conductivity Check
(perchlorate1)

Conductivity Probe Calibration Check: 1380 /1410 µs/cm

Working MCT Level: 10,000 µs/cm

Sample	Conductivity (µs/cm)	Pretreatment or Dilution Needed
WG574101-01 MCT	9,990.	
-02 Blank	0.33	
-03 LCS	0.48	
-04 LCS2	0.51	
L16060823-01	9,120.	
-03	7,440.	
-04	1,943.	
L16060927-01	750.	
-02	652.	
-03	446.	
-04	162.7	
L16061026-01	6,910.	
-03	238.	
-04	965.	
-05	1,526.	
-07	220.	
L16061142-01	165.3	
-02	81.1	
-04	83.1	
-06	728.	
-07	74.2	
L16061144-01	118.6	
-02	118.7	
-03	110.1	

Analyst: John Richards

Date/Time: 06/27/16 / 20:00

DCN#119317

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 050316_JWR.TXT
Analyst1: JWR Analyst2: NA
Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254
Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG567013 (soils)

Internal STD: COA18071 Surrogate STD: NA Calibration STD STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 MS/MSD STD: STD75512

Comments: ICAL WG567320 : Alternate Source STD75512
Analytical Column : RPPX 5um (250x4.6mm)
K'Prime S/N RPPX250-02115

Samples L16041363(-05 and -10) were analyzed at dilutions based on their pre-run screen results.

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	1LM.LM34686	WG567320-01 CCB	1	1		05/03/16 15:06
2	1LM.LM34687	WG567320-02 STD (0.1 ug/L)	1	1	STD75510	05/03/16 15:25
3	1LM.LM34688	WG567320-03 STD (0.2 ug/L)	1	1	STD75510	05/03/16 15:43
4	1LM.LM34689	WG567320-04 STD (0.5 ug/L)	1	1	STD75510	05/03/16 16:02
5	1LM.LM34690	WG567320-05 STD (1.0 ug/L)	1	1	STD75510	05/03/16 16:21
6	1LM.LM34691	WG567320-06 STD (2.0 ug/L)	1	1	STD75510	05/03/16 16:40
7	1LM.LM34692	WG567320-07 STD (5.0 ug/L)	1	1	STD75510	05/03/16 16:59
8	1LM.LM34693	WG567320-08 STD (10 ug/L)	1	1	STD75510	05/03/16 17:18
9	1LM.LM34694	WG567320-09 SSCV (1.0 ug/L)	1	1	STD75512	05/03/16 17:37
10	1LM.LM34695	WG567321-01 CCB	1	1		05/03/16 17:56
11	1LM.LM34696	WG567321-02 CCV (1.0ug/L)	1	1	STD75510	05/03/16 18:15
12	1LM.LM34697	WG567013-07 MRL (2.0ug/kg)	7	1	STD75510	05/03/16 18:34
13	1LM.LM34698	WG567013-01 MCT (2.0ug/kg)	7	1	STD75512	05/03/16 18:53
14	1LM.LM34699	WG567013-02 BLANK	7	1		05/03/16 19:12
15	1LM.LM34700	WG567013-03 LCS (2.0ug/kg)	7	1	STD75512	05/03/16 19:31
16	1LM.LM34701	L16041363-07 RS	7	1		05/03/16 19:50
17	1LM.LM34702	L16041363-08 MS	7	1	STD75512	05/03/16 20:09
18	1LM.LM34703	L16041363-09 MSD	7	1	STD75512	05/03/16 20:28
19	1LM.LM34704	L16041363-01	7	1		05/03/16 20:46
20	1LM.LM34705	L16041363-02	7	1		05/03/16 21:05
21	1LM.LM34706	L16041363-03	7	1		05/03/16 21:24
22	1LM.LM34707	L16041363-04	7	1		05/03/16 21:43
23	1LM.LM34708	WG567321-03 CCV (1.0ug/L)	1	1	STD75510	05/03/16 22:02
24	1LM.LM34709	WG567013-08 MRL (2.0ug/kg)	7	1	STD75510	05/03/16 22:21
25	1LM.LM34710	WG567321-04 CCB	1	1		05/03/16 22:40
26	1LM.LM34711	L16041363-05 (5x)	7	5		05/03/16 22:59
27	1LM.LM34712	L16041363-06	7	1		05/03/16 23:18
28	1LM.LM34713	L16041363-10 (5x)	7	5		05/03/16 23:37
29	1LM.LM34714	WG567321-05 CCV (1.0ug/L)	1	1	STD75510	05/03/16 23:56
30	1LM.LM34715	WG567013-09 MRL (2.0ug/kg)	7	1	STD75510	05/04/16 00:15
31	1LM.LM34716	WG567321-06 CCB	1	1		05/04/16 00:34

Comments

Page: 1 Approved: 05-MAY-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 050316_JWR.TXT

Analyst1: JWR Analyst2: NA

Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254

Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG567013 (soils)

Internal STD: COA18071 Surrogate STD: NA STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 STD75512

Comments

Seq.	Rerun	Dil.	Reason	Analytes
17				
	L16041363-08 MS : The MS %Rec is 129%. This is above the advisory limit of 120%. The parent sample to this MS had responses that passed the ion-ratio criteria, but had a quantified value below the method's detection limit, resulting in an assigned value of zero. If the quantified value for the parent sample were used in the %Rec calculation, the MS %Rec would be 93.9%.			
18				
	L16041363-09 MSD : The MSD %Rec is 131%. This is above the advisory limit of 120%. The parent sample to this MSD had responses that passed the ion-ratio criteria, but had a quantified value below the method's detection limit, resulting in an assigned value of zero. If the quantified value for the parent sample were used in the %Rec calculation, the MSD %Rec would be 95.4%.			

Page: 2 Approved: 05-MAY-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 062416_JWR.TXT
Analyst1: JWR Analyst2: NA
Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254
Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG574101 (waters)

Internal STD: STD18071 Surrogate STD: NA Calibration STD STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 MS/MSD STD: NA

Comments: Samples initially analyzed at dilutions were done so based on their historical results. Samples tagged with (NR) did not conform to their historical results. These samples were reanalyzed on the end of the sequence as neat or at dilutions based on the results of their initial analysis.

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	1LM.LM35462	WG574102-01 CCB	1	1		06/24/16 19:50
2	1LM.LM35463	WG574102-02 CCV (1.0ug/L)	1	1	STD75510	06/24/16 20:09
3	1LM.LM35464	WG574101-05 MRL (0.2ug/L)	1	1	STD75510	06/24/16 20:28
4	1LM.LM35465	WG574101-01 MCT (0.2ug/L)	1	1	STD75512	06/24/16 20:47
5	1LM.LM35466	WG574101-02 BLANK	1	1		06/24/16 21:06
6	1LM.LM35467	WG574101-03 LCS (0.2ug/L)	1	1	STD75512	06/24/16 21:25
7	1LM.LM35468	L16060823-01	1	1		06/24/16 21:44
8	1LM.LM35469	L16060823-03	1	1		06/24/16 22:03
9	1LM.LM35470	L16060823-04 (10x)	1	10		06/24/16 22:22
10	1LM.LM35471	L16060927-01 (100x) (NR)	1	100		06/24/16 22:41
11	1LM.LM35472	L16060927-02 (100x) (NR)	1	100		06/24/16 23:00
12	1LM.LM35473	L16060927-03	1	1		06/24/16 23:18
13	1LM.LM35474	L16060927-04	1	1		06/24/16 23:37
14	1LM.LM35475	WG574102-03 CCV (1.0ug/L)	1	1	STD75510	06/24/16 23:56
15	1LM.LM35476	WG574101-06 MRL (0.2ug/L)	1	1	STD75510	06/25/16 00:15
16	1LM.LM35477	WG574102-04 CCB	1	1		06/25/16 00:34
17	1LM.LM35478	L16061026-01 (100x) (NR)	1	100		06/25/16 00:53
18	1LM.LM35479	L16061026-03 (100x)	1	100		06/25/16 01:12
19	1LM.LM35480	L16061026-04 (10,000x)	1	10000		06/25/16 01:31
20	1LM.LM35481	L16061026-05 (100,000x) (NR)	1	100000		06/25/16 01:50
21	1LM.LM35482	L16061026-07 (10x) (NR)	1	10		06/25/16 02:09
22	1LM.LM35483	L16061142-01	1	1		06/25/16 02:28
23	1LM.LM35484	L16061142-02	1	1		06/25/16 02:47
24	1LM.LM35485	L16061142-04	1	1		06/25/16 03:06
25	1LM.LM35486	L16061142-06 (10x) (NR)	1	10		06/25/16 03:25
26	1LM.LM35487	L16061142-07	1	1		06/25/16 03:44
27	1LM.LM35488	WG574102-05 CCV (1.0ug/L)	1	1	STD75510	06/25/16 04:03
28	1LM.LM35489	WG574101-07 MRL (0.2ug/L)	1	1	STD75510	06/25/16 04:22
29	1LM.LM35490	WG574102-06 CCB	1	1		06/25/16 04:40
30	1LM.LM35491	WG574101-04 LCS2 (0.2ug/L)	1	1	STD75512	06/25/16 04:59
31	1LM.LM35492	L16061144-01	1	1		06/25/16 05:18
32	1LM.LM35493	L16061144-02	1	1		06/25/16 05:37
33	1LM.LM35494	L16061144-03 (10x) (NR)	1	10		06/25/16 05:56

Page: 1 Approved: 28-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 062416_JWR.TXT
Analyst1: JWR Analyst2: NA
Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254
Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG574101 (waters)

Internal STD: STD18071 Surrogate STD: NA STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 NA

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
34	1LM.LM35495	WG574102-07 CCV (1.0ug/L)	1	1	STD75510	06/25/16 06:15
35	1LM.LM35496	WG574101-08 MRL (0.2ug/L)	1	1	STD75510	06/25/16 06:34
36	1LM.LM35497	WG574102-08 CCB	1	1		06/25/16 06:53
37	1LM.LM35498	WG574102-09 CCV (1.0ug/L)	1	1	STD75510	06/25/16 18:06
38	1LM.LM35499	WG574101-09 MRL (0.2ug/L)	1	1	STD75510	06/25/16 18:25
39	1LM.LM35500	WG574102-10 CCB	1	1		06/25/16 18:44
40	1LM.LM35501	L16060927-01 (RR Neat)	1	1		06/25/16 19:02
41	1LM.LM35502	L16060927-02 (RR 1000x)	1	1000		06/25/16 19:21
42	1LM.LM35503	L16061026-01 (RR Neat)	1	1		06/25/16 19:40
43	1LM.LM35504	L16061026-05 (RR 10,000x)	1	10000		06/25/16 19:59
44	1LM.LM35505	L16061026-07 (RR 50x)	1	50		06/25/16 20:18
45	1LM.LM35506	L16061142-06 (RR Neat)	1	1		06/25/16 20:37
46	1LM.LM35507	L16061144-03 (RR Neat)	1	1		06/25/16 20:56
47	1LM.LM35508	WG574102-11 CCV (1.0ug/L)	1	1	STD75510	06/25/16 21:15
48	1LM.LM35509	WG574101-10 MRL (0.2ug/L)	1	1	STD75510	06/25/16 21:34
49	1LM.LM35510	WG574102-12 CCB	1	1		06/25/16 21:53

Comments

Seq.	Rerun	Dil.	Reason	Analytes

Page: 2 Approved: 28-JUN-16

Microbac Laboratories Inc.

Data Checklist

Date: 03-MAY-2016
Analyst: JWR
Analyst: NA
Method: 6850
Instrument: LCMS1
Curve Workgroup: WG567320
Runlog ID: 74891
Analytical Workgroups: L16041363 (SOILS)

ANALYTICAL	
System Performance Check	NA
DFTPP (GCMS)	NA
Endrin/DDT breakdown (8081/GCMS)	NA
Pentachlorophenol/benzidine tailing (GCMS)	NA
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	X
Average RF	NA
Linear regression or higher order curve	X
Alternate source standard (ICV) % Difference	X
Continuing Calibration (CCV)	X
% D/% Drift	X
Minimum response factors (GCMS)	X
Continuing calibration blank (CCB) (IC/LCMS)	X
Limit of quantitation verification (LOQV) (LCMS)	X
Special standards	NA
Blanks	X
TCL hits	ND
Surrogate recoveries	NA
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	NA
MS/MSD/Sample duplicates	X
Recoveries	X
%RPD	X
Interference check sample (ICS) (LCMS)	MCT
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	NA
Surrogate recoveries	NA
Internal standard areas (MS)	X
Library searches (GCMS)	NA
Calculations & correct factors	X
Compounds above calibration range	NA
Reruns	NA
Manual integrations	NA
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	
Check for completeness	X
Primary Reviewer	JWR
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	WTD

Primary Reviewer:
04-MAY-2016
John Richards

Secondary Reviewer:
05-MAY-2016
Wadei D

CHECKLIST1 - Modified 03/05/2008
Generated: MAY-05-2016 16:23:46

Microbac Laboratories Inc.

Data Checklist

Date: 24-JUN-2016
Analyst: JWR
Analyst: NA
Method: 6850
Instrument: LCMS1
Curve Workgroup: NA
Runlog ID: 75960
Analytical Workgroups: L16060823, 0927, 1026, 1142, 1144

ANALYTICAL	
System Performance Check	NA
DFTPP (GCMS)	NA
Endrin/DDT breakdown (8081/GCMS)	NA
Pentachlorophenol/benzidine tailing (GCMS)	NA
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	NA
Average RF	NA
Linear regression or higher order curve	NA
Alternate source standard (ICV) % Difference	NA
Continuing Calibration (CCV)	X
% D/% Drift	X
Minimum response factors (GCMS)	X
Continuing calibration blank (CCB) (IC/LCMS)	X
Limit of quantitation verification (LOQV) (LCMS)	X
Special standards	NA
Blanks	X
TCL hits	ND
Surrogate recoveries	NA
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	NA
MS/MSD/Sample duplicates	NA
Recoveries	NA
%RPD	NA
Interference check sample (ICS) (LCMS)	MCT
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	NA
Surrogate recoveries	NA
Internal standard areas (MS)	X
Library searches (GCMS)	NA
Calculations & correct factors	X
Compounds above calibration range	X
Reruns	X
Manual integrations	MCT-ONLY
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	
Check for completeness	X
Primary Reviewer	JWR
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	WTD

Primary Reviewer:
27-JUN-2016

Secondary Reviewer:
28-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-28-2016 14:00:55

Microbac Laboratories Inc.

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 6850 AAB#: WG574101

Login Number: L16060823

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
126-061416	01	06/14/16					06/24/2016	10.3	28		06/24/16	.1	28	
C08-061416	03	06/14/16					06/24/2016	10.2	28		06/24/16	.1	28	
C03-061416	04	06/14/16					06/24/2016	10.2	28		06/24/16	.1	28	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4827762
Report generated 06/28/2016 14:14

METHOD BLANK SUMMARY

Login Number: L16060823
Blank File ID: 1LM.LM35466
Prep Date: 06/24/16 19:00
Analyzed Date: 06/24/16 21:06
Analyst: JWR

Work Group: WG574101
Blank Sample ID: WG574101-02
Instrument ID: LCMS1
Method: 6850

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
QCMRL	WG574101-05	1LM.LM35464	06/24/16 20:28	01
MCT	WG574101-01	1LM.LM35465	06/24/16 20:47	01
LCS	WG574101-03	1LM.LM35467	06/24/16 21:25	01
126-061416	L16060823-01	1LM.LM35468	06/24/16 21:44	01
C08-061416	L16060823-03	1LM.LM35469	06/24/16 22:03	01
C03-061416	L16060823-04	1LM.LM35470	06/24/16 22:22	DL01
QCMRL	WG574101-06	1LM.LM35476	06/25/16 00:15	01
QCMRL	WG574101-07	1LM.LM35489	06/25/16 04:22	01
LCS2	WG574101-04	1LM.LM35491	06/25/16 04:59	01
QCMRL	WG574101-08	1LM.LM35496	06/25/16 06:34	01
QCMRL	WG574101-09	1LM.LM35499	06/25/16 18:25	01
QCMRL	WG574101-10	1LM.LM35509	06/25/16 21:34	01

Report Name: BLANK_SUMMARY
PDF File ID: 4827763
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060823 Prep Date: 06/24/16 19:00 Sample ID: WG574101-02
Instrument ID: LCMS1 Run Date: 06/24/16 21:06 Prep Method: 6850
File ID: 1LM.LM35466 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Perchlorate	0.100	0.400	0.100	1	U

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name:BLANK
PDF ID: 4827764
28-JUN-2016 14:14

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060823 Analyst: JWR Prep Method: 6850
Instrument ID: LCMS1 Matrix: Water Method: 6850
Workgroup (AAB#): WG574101 Units: ug/L
QC Key: DOD4 Lot #: STD75512

Sample ID: WG574101-03 LCS File ID: 1LM.LM35467 Run Date: 06/24/2016 21:25
Sample ID: WG574101-04 LCS2 File ID: 1LM.LM35491 Run Date: 06/25/2016 04:59

Analytes	LCS			LCS2			%RPD	%Rec Limits	RPD Lmt	Q
	Known	Found	% REC	Known	Found	% REC				
Perchlorate	0.200	0.208	104	0.200	0.213	107	2.38	80 - 120	15	

LCS_LCS2 - Modified 03/06/2008
PDF File ID: 4827765
Report generated: 06/28/2016 14:14

Microbac Laboratories Inc.

INITIAL CALIBRATION SUMMARY

Login Number: L16060823
Analytical Method: 6850
ICAL Workgroup: WG567320

Instrument ID: LCMS1
Initial Calibration Date: 03-MAY-16 17:18
Column ID: F

Analyte		AVG RF	% RSD	LINEAR (R)	QUAD (R^2)
Perchlorate		1.699	4.81	1.00000	

R = Correlation coefficient; 0.995 minimum
R^2 = Coefficient of determination; 0.99 minimum

INT_CAL - Modified 03/06/2008
PDF File ID: 4828694
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16060823

Analytical Method: 6850

Instrument ID: LCMS1

Initial Calibration Date: 03-MAY-16 17:18

Column ID: F

Analyte	WG567320-02			WG567320-03			WG567320-04		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
Perchlorate	0.100	17900.0000	1.792	0.200	34100.0000	1.718	0.500	82200.0000	1.637

INT_CAL - Modified 03/06/2008
PDF File ID: 4828694
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16060823

Analytical Method: 6850

Instrument ID: LCMS1

Initial Calibration Date: 03-MAY-16 17:18

Column ID: F

Analyte	WG567320-05			WG567320-06			WG567320-07		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
Perchlorate	1.00	168000.000	1.697	2.00	330000.000	1.672	5.00	810000.000	1.695

INT_CAL - Modified 03/06/2008
PDF File ID: 4828694
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16060823

Analytical Method: 6850

Instrument ID: LCMS1

Initial Calibration Date: 03-MAY-16 17:18

Column ID: F

Analyte	WG567320-08		
	CONC	RESP	RF
Perchlorate	10.0	1530000.00	1.680

INT_CAL - Modified 03/06/2008
PDF File ID: 4828694
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16060823 Run Date: 05/03/2016 Sample ID: WG567320-09
Instrument ID: LCMS1 Run Time: 17:37 Method: 6850
File ID: 1LM.LM34694 Analyst: JWR QC Key: DOD4
ICal Workgroup: WG567320 Cal ID: LCMS1 - 03-MAY-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
Perchlorate		1.00	0.985	ug/L	1.66	1.50	15	

* Exceeds %D Limit

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4828695
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/24/2016
Sample ID: WG574102-01
Instrument ID: LCMS1
Run Time: 19:50
Method: 6850
File ID: 1LM.LM35462
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827768
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574102-04
Instrument ID: LCMS1
Run Time: 00:34
Method: 6850
File ID: 1LM.LM35477
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827768
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574102-06
Instrument ID: LCMS1
Run Time: 04:40
Method: 6850
File ID: 1LM.LM35490
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827768
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574102-08
Instrument ID: LCMS1
Run Time: 06:53
Method: 6850
File ID: 1LM.LM35497
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827768
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574102-10
Instrument ID: LCMS1
Run Time: 18:44
Method: 6850
File ID: 1LM.LM35500
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827768
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574102-12
Instrument ID: LCMS1
Run Time: 21:53
Method: 6850
File ID: 1LM.LM35510
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827768
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823 Run Date: 06/24/2016 Sample ID: WG574102-02
Instrument ID: LCMS1 Run Time: 20:09 Method: 6850
File ID: 1LM.LM35463 Analyst: JWR QC Key: DOD4
Workgroup (AAB#): WG574101 Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.01	ug/L	1.70	1.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827767
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/24/2016
Sample ID: WG574102-03
Instrument ID: LCMS1
Run Time: 23:56
Method: 6850
File ID: 1LM.LM35475
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.01	ug/L	1.70	1.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827767
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574102-05
Instrument ID: LCMS1
Run Time: 04:03
Method: 6850
File ID: 1LM.LM35488
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.04	ug/L	1.75	4.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827767
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574102-07
Instrument ID: LCMS1
Run Time: 06:15
Method: 6850
File ID: 1LM.LM35495
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.02	ug/L	1.72	2.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827767
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574102-09
Instrument ID: LCMS1
Run Time: 18:06
Method: 6850
File ID: 1LM.LM35498
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.04	ug/L	1.76	4.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827767
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574102-11
Instrument ID: LCMS1
Run Time: 21:15
Method: 6850
File ID: 1LM.LM35508
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.02	ug/L	1.73	2.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827767
Report generated 06/28/2016 14:15

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060823 Run Date: 06/24/2016 Sample ID: WG574101-05

Instrument ID: LCMS1 Run Time: 20:28 Prep Method: 6850

File ID: 1LM.LM35464 Analyst: JWR Method: 6850

Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L

Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.206	103	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827766
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060823 Run Date: 06/25/2016 Sample ID: WG574101-06

Instrument ID: LCMS1 Run Time: 00:15 Prep Method: 6850

File ID: 1LM.LM35476 Analyst: JWR Method: 6850

Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L

Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.198	99.0	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827766
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060823 Run Date: 06/25/2016 Sample ID: WG574101-07
Instrument ID: LCMS1 Run Time: 04:22 Prep Method: 6850
File ID: 1LM.LM35489 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.211	106	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827766
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060823 Run Date: 06/25/2016 Sample ID: WG574101-08
Instrument ID: LCMS1 Run Time: 06:34 Prep Method: 6850
File ID: 1LM.LM35496 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.201	101	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827766
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060823 Run Date: 06/25/2016 Sample ID: WG574101-09

Instrument ID: LCMS1 Run Time: 18:25 Prep Method: 6850

File ID: 1LM.LM35499 Analyst: JWR Method: 6850

Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L

Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.207	104	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827766
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060823 Run Date: 06/25/2016 Sample ID: WG574101-10
Instrument ID: LCMS1 Run Time: 21:34 Prep Method: 6850
File ID: 1LM.LM35509 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.206	103	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827766
Report generated 06/28/2016 14:14

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO AVERAGE OF ICAL)

Login Number: L16060823
Instrument ID: LCMS1
Workgroup (AAB#): WG574101

ICAL CCV Number: WG567320-05
CAL ID: LCMS1-03-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1
WG567320	NA	NA	489000
Upper Limit	NA	NA	733500
Lower Limit	NA	NA	244500
L16060823-01	1.00	01	350000
L16060823-03	1.00	01	343000
L16060823-04	10.0	DL01	405000
WG574101-02	1.00	01	443000
WG574101-03	1.00	01	444000
WG574101-04	1.00	01	526000

IS-1 - O18LP

Underline = Response outside limits

INTERNAL_STD_AVG_ICAL - Modified 03/10/2010
PDF File ID: 4827769
Report generated 06/28/2016 14:15

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	L16060823-01
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35468
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 21:44	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	16900	6640	2.55	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	L16060823-03
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35469
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 22:03	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	6790	3320	2.05	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	L16060823-04
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35470
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 22:22	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	75100	26200	2.87	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG567320-02
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34687
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 15:25	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	17900	6950	2.58	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG567320-03
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34688
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 15:43	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	34100	11900	2.87	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG567320-04
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34689
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:02	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	82200	29400	2.80	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG567320-05
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34690
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:21	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	168000	56600	2.97	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG567320-06
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34691
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:40	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	330000	108000	3.06	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG567320-07
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34692
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:59	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	810000	269000	3.01	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG567320-08
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34693
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 17:18	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	1530000	512000	2.99	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG567320-09
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34694
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 17:37	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	169000	56300	3.00	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	WG574101-01
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35465
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 20:47	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	28100	10000	2.81	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	WG574101-02
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35466
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 21:06	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	WG574101-03
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35467
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 21:25	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	31600	11800	2.68	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	WG574101-04
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35491
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:59	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	38300	14000	2.74	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	WG574101-05
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35464
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 20:28	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	31600	12100	2.61	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	WG574101-06
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35476
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 00:15	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	35000	13100	2.67	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	WG574101-07
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35489
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:22	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	38900	13400	2.90	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	WG574101-08
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35496
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 06:34	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	38000	14100	2.70	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	WG574101-09
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35499
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 18:25	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	32900	11500	2.86	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**	6850	**Samplenum:**	WG574101-10
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35509
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 21:34	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	37900	14000	2.71	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-01
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:50	**File ID:**	1LM.LM35462
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 19:50	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-02
Instrument:	LCMS1	**Prep Date:**	06/24/2016 20:09	**File ID:**	1LM.LM35463
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 20:09	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	152000	55500	2.74	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-03
Instrument:	LCMS1	**Prep Date:**	06/24/2016 23:56	**File ID:**	1LM.LM35475
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 23:56	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	170000	57500	2.96	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-04
Instrument:	LCMS1	**Prep Date:**	06/25/2016 00:34	**File ID:**	1LM.LM35477
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 00:34	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	824	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-05
Instrument:	LCMS1	**Prep Date:**	06/25/2016 04:03	**File ID:**	1LM.LM35488
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:03	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	179000	60200	2.97	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-06
Instrument:	LCMS1	**Prep Date:**	06/25/2016 04:40	**File ID:**	1LM.LM35490
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:40	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-07
Instrument:	LCMS1	**Prep Date:**	06/25/2016 06:15	**File ID:**	1LM.LM35495
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 06:15	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	189000	63400	2.98	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-08
Instrument:	LCMS1	**Prep Date:**	06/25/2016 06:53	**File ID:**	1LM.LM35497
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 06:53	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	518	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-09
Instrument:	LCMS1	**Prep Date:**	06/25/2016 18:06	**File ID:**	1LM.LM35498
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 18:06	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	161000	56700	2.84	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-10
Instrument:	LCMS1	**Prep Date:**	06/25/2016 18:44	**File ID:**	1LM.LM35500
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 18:44	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-11
Instrument:	LCMS1	**Prep Date:**	06/25/2016 21:15	**File ID:**	1LM.LM35508
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 21:15	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	176000	59300	2.97	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060823	**Prep Method:**		**Samplenum:**	WG574102-12
Instrument:	LCMS1	**Prep Date:**	06/25/2016 21:53	**File ID:**	1LM.LM35510
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 21:53	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

2.2.2.3 Sample Data

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Data File	LM35468.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 9:44:17 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16060823-01	**Injection Vial**	8.00
Data File	LM35468.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 9:44:17 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16060823-01	**Dilution Factor**	1.00
Sample Comment	1,1 (Hist)	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	3.500e+05	9.10	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.690e+04	9.10	N/A	0.14
Perchlorate conf	6.640e+03	9.11	N/A	0.149

O18LP (Internal Standard)

RT (Exp. RT): 9.10(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.10 (9.84) min Calculated conc: 0.14 ng/ml Area Ratio: 0.048 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.11 (9.84) min Calculated conc: 0.149 ng/ml Area Ratio: 0.019 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:41 PM

Data File	LM35469.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 10:03:12 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16060823-03	**Injection Vial**	9.00
Data File	LM35469.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 10:03:12 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16060823-03	**Dilution Factor**	1.00
Sample Comment	1,1 (Hist)	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	3.430e+05	9.09	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	6.790e+03	9.12	N/A	0.055
Perchlorate conf	3.320e+03	9.10	N/A	0.0661

O18LP (Internal Standard)

RT (Exp. RT): 9.09(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:41 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.12 (9.84) min
Calculated conc:	0.055 ng/ml
Area Ratio:	0.02
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.10 (9.84) min
Calculated conc:	0.0661 ng/ml
Area Ratio:	0.01
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:41 PM

Data File	LM35470.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 10:22:07 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16060823-04 (10x)	**Injection Vial**	10.00
Data File	LM35470.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 10:22:07 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16060823-04	**Dilution Factor**	1.00
Sample Comment	1,10 (Hist)	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.050e+05	9.49	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	7.510e+04	9.51	N/A	0.548
Perchlorate conf	2.620e+04	9.50	N/A	0.56

O18LP (Internal Standard)

RT (Exp. RT): 9.49(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:41 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.51 (9.84) min
Calculated conc:	0.548 ng/ml
Area Ratio:	0.185
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.50 (9.84) min
Calculated conc:	0.56 ng/ml
Area Ratio:	0.065
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

2.2.2.4 Standards Data

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Data File	LM34686.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-01 CCB	**Injection Vial**	1.00
Data File	LM34686.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-01	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.020e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min
Concentration: 5.00 ug/L
Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:20 PM

Analyte Name: **Perchlorate**
Internal Standard: O18LP

Data File	LM34686.wiff	**Result Table**	050316_JWR.rdb
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Regression Equation: y = 1.68 x + 0.00128 (r = 1.0000)

Expected Concentration	Number of Values	Mean Calculated Concentration	% Accuracy	Std. Deviation	%CV
0.10	1	0.10	102.8	N/A	N/A
0.20	1	0.20	100.3	N/A	N/A
0.50	1	0.48	96.6	N/A	N/A
1.00	1	1.01	100.5	N/A	N/A
2.00	1	1.99	99.3	N/A	N/A
5.00	1	5.04	100.7	N/A	N/A
10.00	1	9.99	99.9	N/A	N/A



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:20 PM

Analyte Name: **Perchlorate conf**
Internal Standard: O18LP

Data File	LM34686.wiff	**Result Table**	050316_JWR.rdb
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Regression Equation: y = 0.558 x + 0.00228 (r = 0.9999)

Expected Concentration	Number of Values	Mean Calculated Concentration	% Accuracy	Std. Deviation	%CV
0.10	1	0.10	104.3	N/A	N/A
0.20	1	0.19	96.8	N/A	N/A
0.50	1	0.50	100.6	N/A	N/A
1.00	1	1.00	100.5	N/A	N/A
2.00	1	1.94	97.2	N/A	N/A
5.00	1	5.02	100.4	N/A	N/A
10.00	1	10.03	100.3	N/A	N/A



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Data File	LM34687.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 3:25:04 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-02 STD (0.1 ug/L)	**Injection Vial**	2.00
Data File	LM34687.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 3:25:04 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-02	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.000e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.790e+04	9.85	0.10	0.103
Perchlorate conf	6.950e+03	9.83	0.10	0.104

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.85 (9.84) min Calculated conc: 0.103 ng/ml Area Ratio: 0.036 Sample Type: (Standard)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.83 (9.84) min Calculated conc: 0.104 ng/ml Area Ratio: 0.014 Sample Type: (Standard)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34688.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 3:43:59 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-03 STD (0.2 ug/L)	**Injection Vial**	3.00
Data File	LM34688.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 3:43:59 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-03	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.970e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.410e+04	9.85	0.20	0.201
Perchlorate conf	1.190e+04	9.84	0.20	0.194

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.201 ng/ml
Area Ratio:	0.069
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	0.194 ng/ml
Area Ratio:	0.024
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34689.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:02:52 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-04 STD (0.5 ug/L)	**Injection Vial**	4.00
Data File	LM34689.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:02:52 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-04	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.020e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	8.220e+04	9.85	0.50	0.483
Perchlorate conf	2.940e+04	9.84	0.50	0.503

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.483 ng/ml
Area Ratio:	0.164
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	0.503 ng/ml
Area Ratio:	0.058
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34690.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:21:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-05 STD (1.0 ug/L)	**Injection Vial**	5.00
Data File	LM34690.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:21:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-05	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.950e+05	9.83	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.680e+05	9.84	1.00	1.01
Perchlorate conf	5.660e+04	9.83	1.00	1.00

O18LP (Internal Standard)

RT (Exp. RT): 9.83(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.339
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	1.00 ng/ml
Area Ratio:	0.114
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34691.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:40:45 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-06 STD (2.0 ug/L)	**Injection Vial**	6.00
Data File	LM34691.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:40:45 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-06	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.930e+05	9.83	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.300e+05	9.84	2.00	1.99
Perchlorate conf	1.080e+05	9.83	2.00	1.94

O18LP (Internal Standard)

RT (Exp. RT): 9.83(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	1.99 ng/ml
Area Ratio:	0.669
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	1.94 ng/ml
Area Ratio:	0.219
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Data File	LM34692.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:59:42 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-07 STD (5.0 ug/L)	**Injection Vial**	7.00
Data File	LM34692.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:59:42 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-07	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.780e+05	9.83	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	8.100e+05	9.84	5.00	5.04
Perchlorate conf	2.690e+05	9.83	5.00	5.02

O18LP (Internal Standard)

RT (Exp. RT): 9.83(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	5.04 ng/ml
Area Ratio:	1.695
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	5.02 ng/ml
Area Ratio:	0.562
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Data File	LM34693.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 5:18:37 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-08 STD (10 ug/L)	**Injection Vial**	8.00
Data File	LM34693.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 5:18:37 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-08	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.560e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.530e+06	9.85	10.00	9.99
Perchlorate conf	5.120e+05	9.83	10.00	10.00

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	9.99 ng/ml
Area Ratio:	3.36
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	10.00 ng/ml
Area Ratio:	1.122
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:23 PM

Data File	LM34694.wiff	**Result Table**	050316_JWR.rdb
Acquisition Date	5/3/2016 5:37:34 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-09 SSCV (1.0 ug/L)	**Injection Vial**	9.00
Data File	LM34694.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 5:37:34 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	050316_JWR.rdb
Sample ID	WG567320-09	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.080e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.690e+05	9.85	1.00	0.985
Perchlorate conf	5.630e+04	9.84	1.00	0.973

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:23 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.985 ng/ml
Area Ratio:	0.332
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	0.973 ng/ml
Area Ratio:	0.111
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Data File	LM35463.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 8:09:38 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-02 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35463.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 8:09:38 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-02	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.450e+05	9.34	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.520e+05	9.35	1.00	1.01
Perchlorate conf	5.550e+04	9.33	1.00	1.10

O18LP (Internal Standard)

RT (Exp. RT): 9.34(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.341
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.33 (9.84) min
Calculated conc:	1.10 ng/ml
Area Ratio:	0.125
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Data File	LM35475.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 11:56:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-03 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35475.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 11:56:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-03	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.980e+05	9.32	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.700e+05	9.33	1.00	1.01
Perchlorate conf	5.750e+04	9.32	1.00	1.01

O18LP (Internal Standard)

RT (Exp. RT): 9.32(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.33 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.34
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.32 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.115
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:47 PM

Data File	LM35488.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:03:05 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-05 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35488.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:03:05 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-05	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.110e+05	9.33	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.790e+05	9.34	1.00	1.04
Perchlorate conf	6.020e+04	9.32	1.00	1.03

O18LP (Internal Standard)

RT (Exp. RT): 9.33(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:47 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.34 (9.84) min
Calculated conc:	1.04 ng/ml
Area Ratio:	0.35
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.32 (9.84) min
Calculated conc:	1.03 ng/ml
Area Ratio:	0.118
Sample Type:	(Quality Control)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:50 PM

Data File	LM35495.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:15:37 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-07 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35495.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:15:37 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-07	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.500e+05	9.33	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.890e+05	9.35	1.00	1.02
Perchlorate conf	6.340e+04	9.33	1.00	1.01

O18LP (Internal Standard)

RT (Exp. RT): 9.33(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:50 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.02 ng/ml
Area Ratio:	0.345
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.33 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.115
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Data File	LM35498.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:06:10 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-09 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35498.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:06:10 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-09	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.580e+05	9.34	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.610e+05	9.35	1.00	1.04
Perchlorate conf	5.670e+04	9.34	1.00	1.09

O18LP (Internal Standard)

RT (Exp. RT): 9.34(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.04 ng/ml
Area Ratio:	0.353
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.34 (9.84) min
Calculated conc:	1.09 ng/ml
Area Ratio:	0.124
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:55 PM

Data File	LM35508.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 9:15:29 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-11 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35508.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 9:15:29 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-11	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.090e+05	9.35	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.760e+05	9.36	1.00	1.02
Perchlorate conf	5.930e+04	9.35	1.00	1.02

O18LP (Internal Standard)

RT (Exp. RT): 9.35(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.36 (9.84) min
Calculated conc:	1.02 ng/ml
Area Ratio:	0.345
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.02 ng/ml
Area Ratio:	0.117
Sample Type:	(Quality Control)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Data File	LM35464.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 8:28:33 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-05 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35464.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 8:28:33 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-05	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.490e+05	9.39	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.160e+04	9.40	N/A	0.206
Perchlorate conf	1.210e+04	9.37	N/A	0.222

O18LP (Internal Standard)

RT (Exp. RT): 9.39(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.40 (9.84) min Calculated conc: 0.206 ng/ml Area Ratio: 0.07 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.37 (9.84) min Calculated conc: 0.222 ng/ml Area Ratio: 0.027 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Data File	LM35476.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 12:15:45 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-06 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35476.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 12:15:45 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-06	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.160e+05	9.36	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.500e+04	9.37	N/A	0.198
Perchlorate conf	1.310e+04	9.35	N/A	0.207

O18LP (Internal Standard)

RT (Exp. RT): 9.36(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.37 (9.84) min
Calculated conc:	0.198 ng/ml
Area Ratio:	0.068
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	0.207 ng/ml
Area Ratio:	0.025
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Data File	LM35489.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:22:00 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-07 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35489.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:22:00 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-07	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.380e+05	9.37	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.890e+04	9.39	N/A	0.211
Perchlorate conf	1.340e+04	9.38	N/A	0.202

O18LP (Internal Standard)

RT (Exp. RT): 9.37(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.39 (9.84) min Calculated conc: 0.211 ng/ml Area Ratio: 0.072 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.38 (9.84) min Calculated conc: 0.202 ng/ml Area Ratio: 0.025 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:50 PM

Data File	LM35496.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:34:33 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-08 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35496.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:34:33 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-08	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.520e+05	9.38	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.800e+04	9.39	N/A	0.201
Perchlorate conf	1.410e+04	9.38	N/A	0.208

O18LP (Internal Standard)

RT (Exp. RT): 9.38(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:50 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.39 (9.84) min
Calculated conc:	0.201 ng/ml
Area Ratio:	0.069
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.38 (9.84) min
Calculated conc:	0.208 ng/ml
Area Ratio:	0.026
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Data File	LM35499.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:25:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-09 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35499.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:25:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-09	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.640e+05	9.39	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.290e+04	9.40	N/A	0.207
Perchlorate conf	1.150e+04	9.38	N/A	0.201

O18LP (Internal Standard)

RT (Exp. RT): 9.39(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.40 (9.84) min
Calculated conc:	0.207 ng/ml
Area Ratio:	0.071
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.38 (9.84) min
Calculated conc:	0.201 ng/ml
Area Ratio:	0.025
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:55 PM

Data File	LM35509.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 9:34:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-10 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35509.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 9:34:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-10	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.370e+05	9.39	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.790e+04	9.40	N/A	0.206
Perchlorate conf	1.400e+04	9.39	N/A	0.214

O18LP (Internal Standard)

RT (Exp. RT): 9.39(9.84) min
Concentration: 5.00 ug/L
Sample Type: (Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:55 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.40 (9.84) min Calculated conc: 0.206 ng/ml Area Ratio: 0.071 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.39 (9.84) min Calculated conc: 0.214 ng/ml Area Ratio: 0.026 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Data File	LM35462.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 7:50:39 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-01 CCB	**Injection Vial**	1.00
Data File	LM35462.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 7:50:39 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-01	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.420e+05	9.58	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.58(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Data File	LM35477.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 12:34:42 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-04 CCB	**Injection Vial**	1.00
Data File	LM35477.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 12:34:42 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-04	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.240e+05	9.57	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	8.240e+02	9.53	N/A	< 0

O18LP (Internal Standard)

RT (Exp. RT): 9.57(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.53 (9.84) min Calculated conc: < 0 ng/ml Area Ratio: 0.002 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Data File	LM35490.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:40:56 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-06 CCB	**Injection Vial**	1.00
Data File	LM35490.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:40:56 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-06	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.420e+05	9.59	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.59(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Data File	LM35497.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:53:28 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-08 CCB	**Injection Vial**	1.00
Data File	LM35497.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:53:28 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-08	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.550e+05	9.59	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	5.180e+02	9.63	N/A	< 0

O18LP (Internal Standard)

RT (Exp. RT): 9.59(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.63 (9.84) min
Calculated conc:	< 0 ng/ml
Area Ratio:	0.001
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Data File	LM35500.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:44:00 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-10 CCB	**Injection Vial**	1.00
Data File	LM35500.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:44:00 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-10	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.840e+05	9.61	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.61(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:56 PM

Data File	LM35510.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 9:53:20 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-12 CCB	**Injection Vial**	1.00
Data File	LM35510.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 9:53:20 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-12	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.380e+05	9.61	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.61(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:56 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Data File	LM35465.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 8:47:28 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-01 MCT (0.2ug/L)	**Injection Vial**	4.00
Data File	LM35465.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 8:47:28 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-01	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.320e+05	8.95	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	2.810e+04	8.97	N/A	0.19
Perchlorate conf	1.000e+04	8.95	N/A	0.188

O18LP (Internal Standard)

RT (Exp. RT): 8.95(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	8.97 (9.84) min
Calculated conc:	0.19 ng/ml
Area Ratio:	0.065
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	8.95 (9.84) min
Calculated conc:	0.188 ng/ml
Area Ratio:	0.023
Sample Type:	(Unknown)



Printing Time: 4:34:28 PM
Printing Date: Saturday, June 25, 2016

Sample Name: "WG574101-01 MCT (0.2ug/L)" Sample ID: "WG574101-01" File: "LM35465.wiff"
Peak Name: "Perchlorate" Mass(es): "98.8/83.3 amu"
Comment: "1,1 STD75512" Annotation: ""

Sample Index:	1
Sample Type:	Unknown
Concentration:	N/A
Calculated Conc:	0.219 ug/L
Acq. Date:	6/24/2016
Acq. Time:	8:47:28 PM
Modified:	No
Proc. Algorithm:	Analyst Classic
Bunching Factor:	2
Noise Threshold:	18.48 cps
Area Threshold:	92.40 cps
Num. Smooths:	5
Exp. Width:	0.20
Exp. Height:	0.01
Exp. Peak Ratio:	5.00
Exp. Adj. Ratio:	4.00
Exp. Val. Ratio:	3.00
RT Window:	30.000 sec
Expected RT:	9.84 min
Use Relative RT:	Yes
Int. Type:	Valley
Retention Time:	8.97 min
Area:	3.23e+004 counts
Height:	1951.523 cps
Start Time:	8.526 min
End Time:	9.417 min



Collected by: N/A
Electronic Signature: no
Operator: lcmsl

Printing Time: 4:36:56 PM
Printing Date: Saturday, June 25, 2016

Sample Name: "WG574101-01 MCT (0.2ug/L)" Sample ID: "WG574101-01" File: "LM35465.wiff"
Peak Name: "Perchlorate" Mass(es): "98.8/83.3 amu"
Comment: "1,1 STD75512" Annotation: ""

Sample Index: 1
Sample Type: Unknown
Concentration: N/A
Calculated Conc: 0.190 ug/L
Acq. Date: 6/24/2016
Acq. Time: 8:47:28 PM

Modified: Yes
Proc. Algorithm: Analyst Classic
Smoothing Factor: 2
Noise Threshold: 18.48 cps
Area Threshold: 92.40 cps
Num. Smooths: 5
Sep. Width: 0.20
Sep. Height: 0.01
Exp. Peak Ratio: 5.00
Exp. Adj. Ratio: 4.00
Exp. Val. Ratio: 3.00
RT Window: 30.000 sec
Expected RT: 9.84 min
Use Relative RT: Yes

Int. Type: Manual
Retention Time: 8.97 min
Area: 2.81e+004 counts
Height: 1872.249 cps
Start Time: 8.587 min
End Time: 9.417 min



#4
JWR 06/25/16

✓ WTD
06-28-16

Collected by: N/A
Electronic Signature: no
Operator: lcms1

2.2.2.5 Raw QC Data

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Data File	LM35466.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 9:06:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-02 BLANK	**Injection Vial**	5.00
Data File	LM35466.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 9:06:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-02	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.430e+05	9.50	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.50(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Data File	LM35467.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 9:25:22 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-03 LCS (0.2ug/L)	**Injection Vial**	6.00
Data File	LM35467.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 9:25:22 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-03	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.440e+05	9.53	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.160e+04	9.54	N/A	0.208
Perchlorate conf	1.180e+04	9.53	N/A	0.217

O18LP (Internal Standard)

RT (Exp. RT): 9.53(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.54 (9.84) min
Calculated conc:	0.208 ng/ml
Area Ratio:	0.071
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.53 (9.84) min
Calculated conc:	0.217 ng/ml
Area Ratio:	0.027
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Data File	LM35491.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:59:54 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-04 LCS2 (0.2ug/L)	**Injection Vial**	7.00
Data File	LM35491.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:59:54 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-04	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.260e+05	9.57	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.830e+04	9.59	N/A	0.213
Perchlorate conf	1.400e+04	9.57	N/A	0.218

O18LP (Internal Standard)

RT (Exp. RT): 9.57(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.59 (9.84) min
Calculated conc:	0.213 ng/ml
Area Ratio:	0.073
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.57 (9.84) min
Calculated conc:	0.218 ng/ml
Area Ratio:	0.027
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

2.3 Metals Data

2.3.1 Metals I C P Data

2.3.1.1 Summary Data

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	126F-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 20:33
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	T3.062316.203318
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	0.100	U	0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Iron, Total	7439-89-6	5.93		0.200	0.100	0.0500
Potassium, Total	7440-09-7	4.36		2.00	1.00	0.500

J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	126F-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/25/2016 12:36
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/25/2016 14:03
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	T3.062516.140359
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	126F-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 21:32
Collect Date:	06/14/2016 11:10	**Dilution:**	20	**File ID:**	T3.062316.213215
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Calcium, Total	7440-70-2	296		10.0	5.00	2.50
Magnesium, Total	7439-95-4	235		20.0	10.0	5.00
Sodium, Total	7440-23-5	829		20.0	10.0	5.00

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C08-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/25/2016 12:36
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/25/2016 14:07
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	T3.062516.140759
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C08-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 20:37
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	T3.062316.203719
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	0.114	J	0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Iron, Total	7439-89-6	0.627		0.200	0.100	0.0500
Potassium, Total	7440-09-7	2.18		2.00	1.00	0.500

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C08-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 21:43
Collect Date:	06/14/2016 13:10	**Dilution:**	20	**File ID:**	T3.062316.214348
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Calcium, Total	7440-70-2	239		10.0	5.00	2.50
Magnesium, Total	7439-95-4	143		20.0	10.0	5.00
Sodium, Total	7440-23-5	761		20.0	10.0	5.00

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C03-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 20:41
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	T3.062316.204123
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	0.100	U	0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Potassium, Total	7440-09-7	4.01		2.00	1.00	0.500

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C03-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/25/2016 12:36
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/25/2016 14:12
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	T3.062516.141200
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	C03-061416	**Prep Method:**	3015	**Prep Date:**	06/17/2016 12:25
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/23/2016 16:45
Workgroup #:	WG573973	**Analyst:**	JYH	**Run Date:**	06/23/2016 21:47
Collect Date:	06/14/2016 14:15	**Dilution:**	20	**File ID:**	T3.062316.214744
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Calcium, Total	7440-70-2	55.4		10.0	5.00	2.50
Iron, Total	7439-89-6	101		4.00	2.00	1.00
Magnesium, Total	7439-95-4	35.8		20.0	10.0	5.00
Sodium, Total	7440-23-5	169		20.0	10.0	5.00

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

2.3.1.2 QC Summary Data

Example 6010 Calculations
Thermo Scientific iCAP

1.0 Initial Calibration (ICAL) Parameters
For a multi-point calibration, the system performs linear regression from data consisting of a blank and four standards.

2.0 Calculating the concentration (C) of an element in water using data from prep log, run log, and quantitation report (note:the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system in ug/mL (ppm)	0.1
Vf = Final volume (mL)	50
Vi = Initial volume (mL)	50
D = Dilution factor as a multiplier (10X = 10)	1
Cx = Concentration of element in ug/mL (mg/L)	0.1

3.0 Calculating the concentration (C) of an element in soil using data from prep log, run log, and quantitation report (note: the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (mg/L) (ppm)	0.1
Vf = Final volume (mL)	50
Vi = Initial weight (g)	1
D = Dilution factor as a multiplier (10X = 10)	1
Cx = Concentration of element in ug/g (mg/kg)	5

4.0 Adjusting the concentration to dry weight:

$$Cdry = \frac{Cx \times 100}{Px}$$

Where:	**Example:**
Cx = Concentration calculated as received (wet basis)	5
Px = Percent solids of sample (%wt)	80
$Cdry$ = Concentration calculated as dry weight (mg/kg)	6.25

Microbac Laboratories Inc.
Microwave Digestion Log

Workgroup: WG573190
Analyst: REK
Spike Analyst: REK
Run Date: 06/17/2016 12:25
Method: 3015
Balance: BAL019
Instrument: MW-4
Instrument Start: 06/17/2016 12:36

SOP: ME407 Revison 19
Spike Solution: STD76576
Spike Witness: VC
HCL Lot #: COA18769
HNO3 Lot #: COA18905
40 & 50 ML. DIGESTION TU COA18987
ICP Filters- fisher-Lot# RGT35619

	SAMPLE #	Type	Matrix	Initial Amount	Final Volume	Initial Vessel Wt	Final Vessel Wt	Spike Amount	Due Date
1	WG573190-03	BLANK	1	40 mL	50 mL	206.067 g	206.054 g		
2	WG573025-01	FBLK1	17	5 mL	50 mL	204.348 g	204.338 g		
3	WG573190-04	LCS	1	40 mL	50 mL	210.758 g	210.752 g	5 mL	
4	WG573190-01	REF	17	5 mL	50 mL	203.898 g	203.882 g		
5	L16060779-02	RS02	17	5 mL	50 mL	203.898 g	203.882 g		06/24/16
6	WG573190-05	MS	17	5 mL	50 mL	207.438 g	207.431 g	5 mL	
7	L16060779-05	MS02	17	5 mL	50 mL	207.438 g	207.431 g	5 mL	06/24/16
8	WG573190-06	MSD	17	5 mL	50 mL	208.146 g	208.126 g	5 mL	
9	L16060779-08	SD02	17	5 mL	50 mL	208.146 g	208.126 g	5 mL	06/24/16
10	L16060779-11	SAMP	17	5 mL	50 mL	203.893 g	203.879 g		06/24/16
11	L16060779-14	SAMP	17	5 mL	50 mL	205.834 g	205.819 g		06/24/16
12	WG573190-02	REF	1	40 mL	50 mL	205.508 g	205.498 g		
13	L16060808-01	RS01	1	40 mL	50 mL	205.508 g	205.498 g		06/24/16
14	WG573190-07	MS	1	40 mL	50 mL	210.145 g	210.129 g	5 mL	
15	L16060808-03	MS01	1	40 mL	50 mL	210.145 g	210.129 g	5 mL	06/24/16
16	WG573190-08	MSD	1	40 mL	50 mL	210.94 g	210.914 g	5 mL	
17	L16060808-04	SD01	1	40 mL	50 mL	210.94 g	210.914 g	5 mL	06/24/16
18	L16060822-02	SAMP	1	40 mL	50 mL	206.366 g	206.337 g		06/24/16
19	L16060822-03	SAMP	1	40 mL	50 mL	204.086 g	204.053 g		06/24/16
20	L16060823-02	SAMP	1	40 mL	50 mL	205.702 g	205.68 g		06/24/16
21	L16060823-03	SAMP	1	40 mL	50 mL	203.815 g	203.795 g		06/24/16
22	L16060823-04	SAMP	1	40 mL	50 mL	205.586 g	205.557 g		06/24/16
23	L16060883-02	SAMP	1	40 mL	50 mL	204.528 g	204.491 g		06/24/16
24	L16060884-01	SAMP	17	5 mL	50 mL	204.107 g	204.076 g		06/22/16
25	L16060896-01	SAMP	1	40 mL	50 mL	205.437 g	205.412 g		06/27/16
26	L16060896-02	SAMP	1	40 mL	50 mL	207.438 g	207.413 g		06/27/16
27	L16060896-03	SAMP	1	40 mL	50 mL	205.762 g	205.735 g		06/27/16
28	L16060896-04	SAMP	1	40 mL	50 mL	205.858 g	205.827 g		06/27/16
29	L16060896-05	SAMP	1	40 mL	50 mL	205.819 g	205.802 g		06/27/16
30	L16060896-07	SAMP	1	40 mL	50 mL	207.197 g	207.165 g		06/27/16

L16060883-02	FILTERED DIGESTATE

Analyst: _____

Reviewer: _____

MW_DIG - Modified 09/30/2009
PDF ID: 4812648
Report generated: 06/17/2016 13:11

TCLP Non-Volatile

Analyst(s): CPD
Date: 6/16/16
Filter Lot #: 9486030
Microbac SOP: TCLP01 Rev #: 12

Analyst / Date		Analyst / Date	
CPD / AMA	6/16/16	CPD	6/17/16
Time On	Temp On °C	Time Off	Temp Off °C
1522	24.6	0814	22.1

Agitator Speed 30 ± 2 rpm ☑

Jug #	Sample #	Tests	Method	Fluid #	Matrix *	% Solid	Pretest pH Initial	Pretest pH Final	Int. Wt. (g)	Fluid Vol. (mL)	Final extract pH
D	06-0779-11	ME	1311	F1-181	S	100	7.01	1.82	25.50	510	4.92
	-02						7.20	1.96	25.35	507	5.01
	-05						7.10	1.78	25.40	508	4.97
	-08						7.13	1.76	25.36	507	4.96
	-14						7.04	1.75	25.35	507	4.98
G-29	06-0829-02	SV		F2-379			12.67	12.62	25.74	514	12.47
D	06-0888-01	300	1312	SPLPF1 -62	S	100	NA	NA	25.12	502	9.47
	-02								25.88	517	9.73
	-03								25.44	508	9.64
	-04								25.43	508	9.75
	-05								25.33	506	9.21
	-06								25.87	517	9.65
	-07								25.52	510	9.60
	-08								25.32	506	9.06
	-09								25.45	509	9.68
	-10								25.27	505	9.97
	-11								25.32	506	9.62
	-12								25.69	513	9.53
	06-0886-01								25.89	517	9.67
	-02								25.16	503	9.55
	-03								25.39	507	9.26
	-04								25.37	507	9.58
	06-0884-01	ME	1311	F1-181	S	100	8.33	3.63	25.22	504	5.61
NA	FBIK1	ME	1311	F1-181	NA	NA	NA	NA	100	100	4.88
	FBIK2	SV		F2-379							2.84
	SPLPF1	300	1312	SPLP F1-62							4.15

CPD 6/16/16

*Matrix Code: (S = solid, sand, soil or sludge) (P = paint) (O = organic) (W = water or aqueous waste)
D = Disposable plastic jug
TCLP Pretest weight will be 5.0 g (± 0.1) unless otherwise noted.
Temperature shall be maintained at 23° ± 2 for 18 ± 2 hours unless otherwise noted.

Comments: N/A

Peer Review By: Allie Alfred 06/17/16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062316T3.2

Analyst1: JYH Analyst2: N/A

Method: 200.7/6010B/6010C SOP: ME600G Rev: 8

Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838

ICSA: STD76382 ICSAB: STD76321 Int. Std: RGT31668

CCV: STD76511 LLCCV: STD76601 Tuning Sol :

Stannous : Hydroxylamine :

Workgroups: 573973,573749,573952

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	T3.062316.162939	WG574093-01	Calibration Point		1		06/23/16 16:29
2	T3.062316.163338	WG574093-02	Calibration Point		1		06/23/16 16:33
3	T3.062316.163737	WG574093-03	Calibration Point		1		06/23/16 16:37
4	T3.062316.164136	WG574093-04	Calibration Point		1		06/23/16 16:41
5	T3.062316.164514	WG574093-05	Calibration Point		1		06/23/16 16:45
6	T3.062316.164851	WG574093-06	Initial Calibration Verification		1		06/23/16 16:48
7	T3.062316.165229	WG574093-07	Initial Calib Blank		1		06/23/16 16:52
8	T3.062316.165617	WG574093-08	Low Level Initial Calibration V		1		06/23/16 16:56
9	T3.062316.170014	WG574093-09	Low Level Initial Calibration V		1		06/23/16 17:00
10	T3.062316.170408	WG574093-10	Interference Check		1		06/23/16 17:04
11	T3.062316.170801	WG574093-11	Interference Check		1		06/23/16 17:08
12	T3.062316.171145	WG574093-12	CCV		1		06/23/16 17:11
13	T3.062316.171523	WG574093-13	CCB		1		06/23/16 17:15
14	T3.062316.171922	WG573484-02	Method/Prep Blank	40/50	1		06/23/16 17:19
15	T3.062316.172321	WG573484-03	Laboratory Control S	40/50	1		06/23/16 17:23
16	T3.062316.172701	WG573484-01	Reference Sample		1	L16060764-04	06/23/16 17:27
17	T3.062316.173057	WG573484-04	Matrix Spike	40/50	1	L16060764-04	06/23/16 17:30
18	T3.062316.173436	WG573484-05	Matrix Spike Duplica	40/50	1	L16060764-04	06/23/16 17:34
19	T3.062316.173816	L16061098-01	5004-G06-WQ-W0266	40/50	1		06/23/16 17:38
20	T3.062316.174215	L16061098-02	5240-G02-WQ-W0087	40/50	1		06/23/16 17:42
21	T3.062316.174614	WG573952-01	Post Digestion Spike		1	L16061098-02	06/23/16 17:46
22	T3.062316.174955	WG573952-02	Serial Dilution		5	L16061098-02	06/23/16 17:49
23	T3.062316.175355	WG573952-02	Serial Dilution		25	L16061098-02	06/23/16 17:53
24	T3.062316.175755	WG574093-14	CCV		1		06/23/16 17:57
25	T3.062316.180133	WG574093-15	CCB		1		06/23/16 18:01
26	T3.062316.180532	L16061098-03	5320-G02-WQ-W0183	40/50	1		06/23/16 18:05
27	T3.062316.180930	L16061098-04	5506-G02-WQ-W0164	40/50	1		06/23/16 18:09
28	T3.062316.181329	L16060764-01	MW-23-004	40/50	1		06/23/16 18:13
29	T3.062316.181726	L16060764-02	MW-23-002	40/50	1		06/23/16 18:17
30	T3.062316.182122	L16060764-03	FD-061316	40/50	1		06/23/16 18:21
31	T3.062316.182520	L16060957-01	15-12-13 S7	40/50	1		06/23/16 18:25
32	T3.062316.182915	L16060957-02	15-12-13 S7	40/50	1		06/23/16 18:29
33	T3.062316.183311	WG574093-16	CCV		1		06/23/16 18:33
34	T3.062316.183649	WG574093-17	CCB		1		06/23/16 18:36

Page: 1 Approved: June 26, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062316T3.2
Analyst1: JYH Analyst2: N/A
Method: 200.7/6010B/6010C SOP: ME600G Rev: 8
Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838
ICSA: STD76382 ICSAB: STD76321 Int. Std: RGT31668
CCV: STD76511 LLCCV: STD76601 Tuning Sol :
Stannous : Hydroxylamine :

Workgroups: 573973,573749,573952

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	T3.062316.184048	L16060957-03	15-12-13 S8	40/50	1		06/23/16 18:40
36	T3.062316.184442	L16060957-04	15-12-13 S8	40/50	1		06/23/16 18:44
37	T3.062316.184837	L16060957-05	15-12-13 S5	40/50	1		06/23/16 18:48
38	T3.062316.185233	L16060961-01	GH58HS_MELZER_03-03-0	40/50	1		06/23/16 18:52
39	T3.062316.185630	L16060961-02	GH58HS_MELZER_03-03-0	40/50	1		06/23/16 18:56
40	T3.062316.190026	L16060991-02	MW31-061616		1		06/23/16 19:00
41	T3.062316.190423	L16060995-01	UNTREATED	40/50	1		06/23/16 19:04
42	T3.062316.190829	L16060995-02	AFTER	40/50	1		06/23/16 19:08
43	T3.062316.191235	L16061064-01	59-8-6.02 W1	40/50	1		06/23/16 19:12
44	T3.062316.191630	WG574093-18	CCV		1		06/23/16 19:16
45	T3.062316.192008	WG574093-19	CCB		1		06/23/16 19:20
46	T3.062316.192408	WG573190-03	Method/Prep Blank	40/50	1		06/23/16 19:24
47	T3.062316.192807	WG573190-04	Laboratory Control S	40/50	1		06/23/16 19:28
48	T3.062316.193148	WG573025-01	Fluid Blank 1		1		06/23/16 19:31
49	T3.062316.193546	WG573190-01	Reference Sample		1	L16060779-02	06/23/16 19:35
50	T3.062316.193944	WG573190-05	Matrix Spike	5/50	1	L16060779-02	06/23/16 19:39
51	T3.062316.194323	WG573190-06	Matrix Spike Duplica	5/50	1	L16060779-02	06/23/16 19:43
52	T3.062316.194701	L16060779-11	477-TS02-C0002	5/50	1		06/23/16 19:47
53	T3.062316.195059	WG573973-01	Post Digestion Spike		1	L16060779-11	06/23/16 19:50
54	T3.062316.195438	WG573973-02	Serial Dilution		5	L16060779-11	06/23/16 19:54
55	T3.062316.195838	WG573973-02	Serial Dilution		25	L16060779-11	06/23/16 19:58
56	T3.062316.200238	WG574093-20	CCV		1		06/23/16 20:02
57	T3.062316.200616	WG574093-21	CCB		1		06/23/16 20:06
58	T3.062316.201014	L16060779-14	477-TS01-C0003	5/50	1		06/23/16 20:10
59	T3.062316.201412	WG573190-02	Reference Sample		1	L16060808-01	06/23/16 20:14
60	T3.062316.201808	WG573190-07	Matrix Spike	40/50	1	L16060808-01	06/23/16 20:18
61	T3.062316.202147	WG573190-08	Matrix Spike Duplica	40/50	1	L16060808-01	06/23/16 20:21
62	T3.062316.202527	L16060822-02	18WW03F-061416	40/50	1		06/23/16 20:25
63	T3.062316.202923	L16060822-03	18CPTMW04SW-061416	40/50	1		06/23/16 20:29
64	T3.062316.203318	L16060823-02	126F-061416	40/50	1		06/23/16 20:33
65	T3.062316.203719	L16060823-03	C08-061416	40/50	1		06/23/16 20:37
66	T3.062316.204123	L16060823-04	C03-061416	40/50	1		06/23/16 20:41
67	T3.062316.204516	L16060883-02	WC02	40/50	1		06/23/16 20:45
68	T3.062316.204910	WG574093-22	CCV		1		06/23/16 20:49

Page: 2 Approved: June 26, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062316T3.2

Analyst1: JYH Analyst2: N/A

Method: 200.7/6010B/6010C SOP: ME600G Rev: 8

Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838

ICSA: STD76382 ICSAB: STD76321 Int. Std: RGT31668

CCV: STD76511 LLCCV: STD76601 Tuning Sol :

Stannous : Hydroxylamine :

Workgroups: 573973,573749,573952

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
69	T3.062316.205247	WG574093-23	CCB		1		06/23/16 20:52
70	T3.062316.205645	L16060884-01	NI CAKE	5/50	1		06/23/16 20:56
71	T3.062316.210040	L16060896-01	A08-MW02-Y1S2	40/50	1		06/23/16 21:00
72	T3.062316.210433	L16060896-02	A08-MW03-Y1S2	40/50	1		06/23/16 21:04
73	T3.062316.210831	L16060896-03	A08-MW10-Y1S2	40/50	1		06/23/16 21:08
74	T3.062316.211228	L16060896-04	FD-01	40/50	1		06/23/16 21:12
75	T3.062316.211622	L16060896-05	FD-03	40/50	1		06/23/16 21:16
76	T3.062316.212019	L16060896-07	RB01-06-15-16	40/50	1		06/23/16 21:20
77	T3.062316.212418	L16060822-02	18WW03F-061416		20		06/23/16 21:24
78	T3.062316.212816	L16060822-03	18CPTMW04SW-061416	40/50	20		06/23/16 21:28
79	T3.062316.213215	L16060823-02	126F-061416	40/50	20		06/23/16 21:32
80	T3.062316.213611	WG574093-24	CCV		1		06/23/16 21:36
81	T3.062316.213949	WG574093-25	CCB		1		06/23/16 21:39
82	T3.062316.214348	L16060823-03	C08-061416	40/50	20		06/23/16 21:43
83	T3.062316.214744	L16060823-04	C03-061416	40/50	20		06/23/16 21:47
84	T3.062316.215141	L16060883-02	WC02	40/50	20		06/23/16 21:51
85	T3.062316.215540	L16060896-01	A08-MW02-Y1S2	40/50	20		06/23/16 21:55
86	T3.062316.215938	L16060896-02	A08-MW03-Y1S2	40/50	20		06/23/16 21:59
87	T3.062316.220337	L16060896-03	A08-MW10-Y1S2	40/50	20		06/23/16 22:03
88	T3.062316.220736	L16060896-04	FD-01	40/50	20		06/23/16 22:07
89	T3.062316.221134	L16060896-05	FD-03	40/50	20		06/23/16 22:11
90	T3.062316.221533	L16060896-07	RB01-06-15-16	40/50	20		06/23/16 22:15
91	T3.062316.221931	WG574093-26	CCV		1		06/23/16 22:19
92	T3.062316.222310	WG574093-27	CCB		1		06/23/16 22:23
93	T3.062316.222708	WG574093-28	Low Level Continuing Calibra		1		06/23/16 22:27
94	T3.062316.223105	WG574093-29	Low Level Continuing Calibra		1		06/23/16 22:31
95	T3.062316.223500	WG573562-02	Method/Prep Blank	40/50	1		06/23/16 22:35
96	T3.062316.223859	WG573562-03	Laboratory Control S	40/50	1		06/23/16 22:38
97	T3.062316.224239	WG573353-01	Fluid Blank 1		1		06/23/16 22:42
98	T3.062316.224638	WG573353-02	Fluid Blank 2		1		06/23/16 22:46
99	T3.062316.225037	WG573562-01	Reference Sample		1	L16060974-01	06/23/16 22:50
100	T3.062316.225432	WG573562-04	Matrix Spike	40/50	1	L16060974-01	06/23/16 22:54
101	T3.062316.225812	WG573562-05	Matrix Spike Duplica	40/50	1	L16060974-01	06/23/16 22:58
102	T3.062316.230151	L16060903-01	LF30/31-EWTS-EFFLUENT-1	40/50	1		06/23/16 23:01

Page: 3 Approved: June 26, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062316T3.2

Analyst1: JYH Analyst2: N/A

Method: 200.7/6010B/6010C SOP: ME600G Rev: 8

Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838

ICSA: STD76382 ICSAB: STD76321 Int. Std: RGT31668

CCV: STD76511 LLCCV: STD76601 Tuning Sol :

Stannous : Hydroxylamine :

Workgroups: 573973,573749,573952

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
103	T3.062316.230549	WG573749-03	Post Digestion Spike		1	L16060903-01	06/23/16 23:05
104	T3.062316.230928	WG573749-04	Serial Dilution		5	L16060903-01	06/23/16 23:09
105	T3.062316.231327	WG574093-30	CCV		1		06/23/16 23:13
106	T3.062316.231705	WG574093-31	CCB		1		06/23/16 23:17
107	T3.062316.232104	L16060903-02	LF30/31-EWTS-MH3-15JUNI	40/50	1		06/23/16 23:21
108	T3.062316.232500	L16060903-03	LF30/31-EWTS-MH2-15JUNI	40/50	1		06/23/16 23:25
109	T3.062316.232857	L16060903-04	LF30/31-EWTS-MH1-15JUNI	40/50	1		06/23/16 23:28
110	T3.062316.233254	L16060903-05	LF30/31-EWTS-INFLUENT-1	40/50	1		06/23/16 23:32
111	T3.062316.233650	L16060924-01	K6F0495-01	5/50	1		06/23/16 23:36
112	T3.062316.234055	DI7	DI7		20		06/23/16 23:40
113	T3.062316.234454	DI8	DI8		20		06/23/16 23:44
114	T3.062316.234853	DI9	DI9		20		06/23/16 23:48
115	T3.062316.235252	WG574093-32	CCV		1		06/23/16 23:52
116	T3.062316.235630	WG574093-33	CCB		1		06/23/16 23:56
117	T3.062416.000030	WG574093-34	ICSA		1		06/24/16 00:00
118	T3.062416.000429	WG574093-35	ICSAB		1		06/24/16 00:04
119	T3.062416.000822	WG574093-36	CCV		1		06/24/16 00:08
120	T3.062416.001200	WG574093-37	CCB		1		06/24/16 00:12
121	T3.062416.001600	WG574093-38	Low Level Continuing Calibra		1		06/24/16 00:16
122	T3.062416.001957	WG574093-39	Low Level Continuing Calibra		1		06/24/16 00:19

Page: 4 Approved: June 26, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062516T3.1R.TXT
Analyst1: JYH Analyst2: N/A
Method: 200.7/6010B/6010C SOP: ME600G Rev: 8
Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838
ICSA: STD76382 ICSAB: STD76834 Int. Std: RGT31668
CCV: STD76511 LLCCV: STD76601 Tuning Sol :
Stannous : Hydroxylamine :

Workgroups: 573973,574055,574076,574216,574215,573952

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	T3.062516.122104	WG574249-01	Calibration Point		1		06/25/16 12:21
2	T3.062516.122502	WG574249-02	Calibration Point		1		06/25/16 12:25
3	T3.062516.122900	WG574249-03	Calibration Point		1		06/25/16 12:29
4	T3.062516.123256	WG574249-04	Calibration Point		1		06/25/16 12:32
5	T3.062516.123634	WG574249-05	Calibration Point		1		06/25/16 12:36
6	T3.062516.124009	WG574249-06	Initial Calibration Verification		1		06/25/16 12:40
7	T3.062516.124348	WG574249-07	Initial Calib Blank		1		06/25/16 12:43
8	T3.062516.124745	WG574249-08	Low Level Initial Calibration V		1		06/25/16 12:47
9	T3.062516.125139	WG574249-09	Low Level Initial Calibration V		1		06/25/16 12:51
10	T3.062516.125529	WG574249-10	Interference Check		1		06/25/16 12:55
11	T3.062516.125921	WG574249-11	Interference Check		1		06/25/16 12:59
12	T3.062516.130304	WG574249-12	CCV		1		06/25/16 13:03
13	T3.062516.130641	WG574249-13	CCB		1		06/25/16 13:06
14	T3.062516.131816	WG573190-03	Method/Prep Blank	40/50	1		06/25/16 13:18
15	T3.062516.132214	WG573190-04	Laboratory Control S	40/50	1		06/25/16 13:22
16	T3.062516.132554	WG573190-02	Reference Sample		1	L16060808-01	06/25/16 13:25
17	T3.062516.132950	WG573190-07	Matrix Spike	40/50	1	L16060808-01	06/25/16 13:29
18	T3.062516.133327	WG573190-08	Matrix Spike Duplica	40/50	1	L16060808-01	06/25/16 13:33
19	T3.062516.133704	L16060822-03	18CPTMW04SW-061416	40/50	1		06/25/16 13:37
20	T3.062516.134058	L16060822-02	18WW03F-061416	40/50	1		06/25/16 13:40
21	T3.062516.134452	WG573973-03	Post Digestion Spike		1	L16060822-02	06/25/16 13:44
22	T3.062516.134830	WG573973-04	Serial Dilution		5	L16060822-02	06/25/16 13:48
23	T3.062516.135227	WG573973-04	Serial Dilution		25	L16060822-02	06/25/16 13:52
24	T3.062516.135622	WG574249-14	CCV		1		06/25/16 13:56
25	T3.062516.140001	WG574249-15	CCB		1		06/25/16 14:00
26	T3.062516.140359	L16060823-02	126F-061416	40/50	1		06/25/16 14:03
27	T3.062516.140759	L16060823-03	C08-061416	40/50	1		06/25/16 14:07
28	T3.062516.141200	L16060823-04	C03-061416	40/50	1		06/25/16 14:12
29	T3.062516.141551	L16061148-01	DP005SW0513-007	40/50	1		06/25/16 14:15
30	T3.062516.141945	L16061148-02	DP005SW0514-007	40/50	1		06/25/16 14:19
31	T3.062516.142340	L16061150-01	SS007SW0709-007	40/50	1		06/25/16 14:23
32	T3.062516.142732	L16061150-02	SS007SW0710-007	40/50	1		06/25/16 14:27
33	T3.062516.143125	L16061151-02	LF003SW0311-007 DUP	40/50	1		06/25/16 14:31
34	T3.062516.143519	L16061151-05	LF003SW0312-007	40/50	1		06/25/16 14:35

Page: 1 Approved: June 27, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062516T3.1R.TXT
Analyst1: JYH Analyst2: N/A
Method: 200.7/6010B/6010C SOP: ME600G Rev: 8
Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838
ICSA: STD76382 ICSAB: STD76834 Int. Std: RGT31668
CCV: STD76511 LLCCV: STD76601 Tuning Sol :
Stannous : Hydroxylamine :

Workgroups: 573973,574055,574076,574216,574215,573952

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	T3.062516.143913	WG573707-02	Reference Sample		1	L16061154-01	06/25/16 14:39
36	T3.062516.144313	WG574249-16	CCV		1		06/25/16 14:43
37	T3.062516.144639	WG574249-17	CCB		1		06/25/16 14:46
38	T3.062516.145037	WG573707-07	Matrix Spike	40/50	1	L16061154-01	06/25/16 14:50
39	T3.062516.145424	WG573707-08	Matrix Spike Duplica	40/50	1	L16061154-01	06/25/16 14:54
40	T3.062516.145809	L16061154-04	FQC-RB06201601-CB	40/50	1		06/25/16 14:58
41	T3.062516.150207	L16061155-01	ST008MW0812-007	40/50	1		06/25/16 15:02
42	T3.062516.150559	L16061155-02	ST008MW0809-007	40/50	1		06/25/16 15:05
43	T3.062516.150954	L16061155-03	ST008MW0816-007	40/50	1		06/25/16 15:09
44	T3.062516.151348	L16061178-01	2016PRE-7217NEWENGLAN	40/50	1		06/25/16 15:13
45	T3.062516.151743	L16061227-01	18WW22-062116	40/50	20		06/25/16 15:17
46	T3.062516.152125	L16061227-06	18CPTMW18-062116	40/50	20		06/25/16 15:21
47	T3.062516.152518	WG573994-01	Reference Sample		20	L16061279-02	06/25/16 15:25
48	T3.062516.152911	WG574249-18	CCV		1		06/25/16 15:29
49	T3.062516.153248	WG574249-19	CCB		1		06/25/16 15:32
50	T3.062516.153646	WG573994-04	Matrix Spike	40/50	20	L16061279-02	06/25/16 15:36
51	T3.062516.154039	WG573994-05	Matrix Spike Duplica	40/50	20	L16061279-02	06/25/16 15:40
52	T3.062516.154432	L16061289-01	MW21-062216	40/50	20		06/25/16 15:44
53	T3.062516.154826	L16061129-01	MW-22-002		20		06/25/16 15:48
54	T3.062516.155222	L16061129-02	MW-22-010		20		06/25/16 15:52
55	T3.062516.155620	L16061129-03	MW-22-007		20		06/25/16 15:56
56	T3.062516.160018	L16061129-04	FD061716		20		06/25/16 16:00
57	T3.062516.160415	L16061272-01	SS007MW0B18-007		20		06/25/16 16:04
58	T3.062516.160812	L16061272-02	SS007MWN752-007		20		06/25/16 16:08
59	T3.062516.161208	L16061279-01	XU936TW007A	40/50	20		06/25/16 16:12
60	T3.062516.161603	WG574249-20	CCV		1		06/25/16 16:16
61	T3.062516.161940	WG574249-21	CCB		1		06/25/16 16:19
62	T3.062516.162338	L16061279-05	XU936TW019A	40/50	20		06/25/16 16:23
63	T3.062516.162733	L16061148-01	DP005SW0513-007		20		06/25/16 16:27
64	T3.062516.163130	L16061148-02	DP005SW0514-007		20		06/25/16 16:31
65	T3.062516.163527	L16061150-01	SS007SW0709-007		20		06/25/16 16:35
66	T3.062516.163912	L16061151-02	LF003SW0311-007 DUP		20		06/25/16 16:39
67	T3.062516.164310	L16061151-05	LF003SW0312-007		20		06/25/16 16:43
68	T3.062516.164706	WG573707-02	Reference Sample		20	L16061154-01	06/25/16 16:47

Page: 2 Approved: June 27, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062516T3.1R.TXT
Analyst1: JYH Analyst2: N/A
Method: 200.7/6010B/6010C SOP: ME600G Rev: 8
Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838
ICSA: STD76382 ICSAB: STD76834 Int. Std: RGT31668
CCV: STD76511 LLCCV: STD76601 Tuning Sol :
Stannous : Hydroxylamine :

Workgroups: 573973,574055,574076,574216,574215,573952

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
69	T3.062516.165101	WG573707-07	Matrix Spike	40/50	20	L16061154-01	06/25/16 16:51
70	T3.062516.165454	L16061279-08	XU936TW011MSD	40/50	20	WG573707-08	06/25/16 16:54
71	T3.062516.165848	WG574249-22	CCV		1		06/25/16 16:58
72	T3.062516.170225	WG574249-23	CCB		1		06/25/16 17:02
73	T3.062516.170624	L16061155-01	ST008MW0812-007		20		06/25/16 17:06
74	T3.062516.171010	L16061155-02	ST008MW0809-007		20		06/25/16 17:10
75	T3.062516.171407	L16061155-03	ST008MW0816-007		20		06/25/16 17:14
76	T3.062516.171804	WG574249-24	CCV		1		06/25/16 17:18
77	T3.062516.172142	WG574249-25	CCB		1		06/25/16 17:21
78	T3.062516.172540	WG574249-26	Low Level Continuing Calibra		1		06/25/16 17:25
79	T3.062516.172933	WG573530-03	Method/Prep Blank	40/50	1		06/25/16 17:29
80	T3.062516.173331	WG573530-04	Laboratory Control S	40/50	1		06/25/16 17:33
81	T3.062516.173711	WG573530-05	Filter Blank		1		06/25/16 17:37
82	T3.062516.174110	L16060986-02	MW-07-04	40/50	1		06/25/16 17:41
83	T3.062516.174505	L16060986-04	MW-01-10	40/50	1		06/25/16 17:45
84	T3.062516.174856	L16060986-06	MW-03-12	40/50	1		06/25/16 17:48
85	T3.062516.175251	L16060986-08	MW-01-07	40/50	1		06/25/16 17:52
86	T3.062516.175646	WG574216-01	Post Digestion Spike		1	L16060986-08	06/25/16 17:56
87	T3.062516.180025	WG574216-02	Serial Dilution		5	L16060986-08	06/25/16 18:00
88	T3.062516.180418	WG574216-02	Serial Dilution		25	L16060986-08	06/25/16 18:04
89	T3.062516.180815	WG574249-27	CCV		1		06/25/16 18:08
90	T3.062516.181152	WG574249-28	CCB		1		06/25/16 18:11
91	T3.062516.181548	L16060986-10	MW-03-11		1		06/25/16 18:15
92	T3.062516.181942	L16060986-12	MW-01-01		1		06/25/16 18:19
93	T3.062516.182332	L16060986-14	MW-03-09		1		06/25/16 18:23
94	T3.062516.182724	L16060986-16	MW-01-08		1		06/25/16 18:27
95	T3.062516.183114	L16060986-18	MW-03-06		1		06/25/16 18:31
96	T3.062516.183511	L16060986-20	MW-03-05		1		06/25/16 18:35
97	T3.062516.183905	WG573530-01	Reference Sample		1	L16060986-22	06/25/16 18:39
98	T3.062516.184258	WG573530-02	Reference Sample		1	L16060986-24	06/25/16 18:42
99	T3.062516.184649	L16060986-26	MW-01-02		1		06/25/16 18:46
100	T3.062516.185044	L16060986-28	DUP 1		1		06/25/16 18:50
101	T3.062516.185437	WG574249-29	CCV		1		06/25/16 18:54
102	T3.062516.185815	WG574249-30	CCB		1		06/25/16 18:58

Page: 3 Approved: June 27, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062516T3.1R.TXT
Analyst1: JYH Analyst2: N/A
Method: 200.7/6010B/6010C SOP: ME600G Rev: 8
Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838
ICSA: STD76382 ICSAB: STD76834 Int. Std: RGT31668
CCV: STD76511 LLCCV: STD76601 Tuning Sol :
Stannous : Hydroxylamine :

Workgroups: 573973,574055,574076,574216,574215,573952

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
103	T3.062516.190211	L16060986-30	DUP 2		1		06/25/16 19:02
104	T3.062516.190604	WG573530-06	Matrix Spike		1	L16060986-22	06/25/16 19:06
105	T3.062516.190942	WG573530-07	Matrix Spike Duplica		1	L16060986-22	06/25/16 19:09
106	T3.062516.191321	WG573530-08	Matrix Spike		1	L16060986-24	06/25/16 19:13
107	T3.062516.191658	WG573530-09	Matrix Spike Duplica		1	L16060986-24	06/25/16 19:16
108	T3.062516.192035	L16060986-40	EQUIP BLANK 1		1		06/25/16 19:20
109	T3.062516.192431	L16060986-42	EQUIP BLANK 2		1		06/25/16 19:24
110	T3.062516.192827	WG574249-31	CCV		1		06/25/16 19:28
111	T3.062516.193205	WG574249-32	CCB		1		06/25/16 19:32
112	T3.062516.193603	WG573544-03	Method/Prep Blank	40/50	1		06/25/16 19:36
113	T3.062516.193959	WG573544-04	Laboratory Control S	40/50	1		06/25/16 19:39
114	T3.062516.194338	L16060907-23	MW36-061516	40/50	1		06/25/16 19:43
115	T3.062516.194731	L16060986-01	MW-07-04	40/50	1		06/25/16 19:47
116	T3.062516.195124	L16060986-03	MW-01-10	40/50	1		06/25/16 19:51
117	T3.062516.195515	L16060986-05	MW-03-12	40/50	1		06/25/16 19:55
118	T3.062516.195905	L16060986-07	MW-01-07	40/50	1		06/25/16 19:59
119	T3.062516.200300	WG574215-01	Post Digestion Spike		1	L16060986-07	06/25/16 20:03
120	T3.062516.200637	WG574215-02	Serial Dilution		5	L16060986-07	06/25/16 20:06
121	T3.062516.201030	WG574215-02	Serial Dilution		25	L16060986-07	06/25/16 20:10
122	T3.062516.201426	WG574249-33	CCV		1		06/25/16 20:14
123	T3.062516.201803	WG574249-34	CCB		1		06/25/16 20:18
124	T3.062516.202200	L16060986-09	MW-03-11	40/50	1		06/25/16 20:22
125	T3.062516.202553	L16060986-11	MW-01-01	40/50	1		06/25/16 20:25
126	T3.062516.202942	L16060986-13	MW-03-09	40/50	1		06/25/16 20:29
127	T3.062516.203332	L16060986-15	MW-01-08	40/50	1		06/25/16 20:33
128	T3.062516.203723	L16060986-17	MW-03-06	40/50	1		06/25/16 20:37
129	T3.062516.204116	L16060986-19	MW-03-05	40/50	1		06/25/16 20:41
130	T3.062516.204510	WG573544-01	Reference Sample		1	L16060986-21	06/25/16 20:45
131	T3.062516.204903	WG573544-02	Reference Sample		1	L16060986-23	06/25/16 20:49
132	T3.062516.205255	L16060986-25	MW-01-02	40/50	1		06/25/16 20:52
133	T3.062516.205646	L16060986-27	DUP 1	40/50	1		06/25/16 20:56
134	T3.062516.210036	WG574249-35	CCV		1		06/25/16 21:00
135	T3.062516.210414	WG574249-36	CCB		1		06/25/16 21:04
136	T3.062516.210812	L16060986-29	DUP 2	40/50	1		06/25/16 21:08

Page: 4 Approved: June 27, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062516T3.1R.TXT
Analyst1: JYH Analyst2: N/A
Method: 200.7/6010B/6010C SOP: ME600G Rev: 8
Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838
ICSA: STD76382 ICSAB: STD76834 Int. Std: RGT31668
CCV: STD76511 LLCCV: STD76601 Tuning Sol :
Stannous : Hydroxylamine :

Workgroups: 573973,574055,574076,574216,574215,573952

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
137	T3.062516.211202	WG573544-05	Matrix Spike	40/50	1	L16060986-21	06/25/16 21:12
138	T3.062516.211541	WG573544-06	Matrix Spike Duplica	40/50	1	L16060986-21	06/25/16 21:15
139	T3.062516.211919	WG573544-07	Matrix Spike	40/50	1	L16060986-23	06/25/16 21:19
140	T3.062516.212256	WG573544-08	Matrix Spike Duplica	40/50	1	L16060986-23	06/25/16 21:22
141	T3.062516.212633	L16060986-39	EQUIP BLANK 1	40/50	1		06/25/16 21:26
142	T3.062516.213031	L16060986-41	EQUIP BLANK 2	40/50	1		06/25/16 21:30
143	T3.062516.213430	WG574249-37	CCV		1		06/25/16 21:34
144	T3.062516.213807	WG574249-38	CCB		1		06/25/16 21:38
145	T3.062516.214203	WG573102-02	Method/Prep Blank	40/50	1		06/25/16 21:42
146	T3.062516.214559	WG573102-03	Laboratory Control S	40/50	1		06/25/16 21:45
147	T3.062516.214937	WG573102-04	Filter Blank		1		06/25/16 21:49
148	T3.062516.215335	WG573102-01	Reference Sample		1	L16060787-09	06/25/16 21:53
149	T3.062516.215728	WG573102-05	Matrix Spike	40/50	1	L16060787-09	06/25/16 21:57
150	T3.062516.220106	WG573102-06	Matrix Spike Duplica	40/50	1	L16060787-09	06/25/16 22:01
151	T3.062516.220444	L16060787-01	EB04	40/50	1		06/25/16 22:04
152	T3.062516.220842	WG573561-03	Post Digestion Spike		1	L16060787-01	06/25/16 22:08
153	T3.062516.221220	WG573561-04	Serial Dilution		5	L16060787-01	06/25/16 22:12
154	T3.062516.221616	WG573561-04	Serial Dilution		25	L16060787-01	06/25/16 22:16
155	T3.062516.222012	WG574249-39	CCV		1		06/25/16 22:20
156	T3.062516.222349	WG574249-40	CCB		1		06/25/16 22:23
157	T3.062516.222746	L16060787-10	U517TW006C	40/50	1		06/25/16 22:27
158	T3.062516.223140	L16060787-18	U517TW008A	40/50	1		06/25/16 22:31
159	T3.062516.223534	L16060787-19	EB05	40/50	1		06/25/16 22:35
160	T3.062516.223931	WG573102-01	Reference Sample		20	L16060787-09	06/25/16 22:39
161	T3.062516.224326	WG573102-05	Matrix Spike	40/50	20	L16060787-09	06/25/16 22:43
162	T3.062516.224719	WG573102-06	Matrix Spike Duplica	40/50	20	L16060787-09	06/25/16 22:47
163	T3.062516.225113	WG574249-41	CCV		1		06/25/16 22:51
164	T3.062516.225451	WG574249-42	CCB		1		06/25/16 22:54
165	T3.062516.225848	WG573484-02	Method/Prep Blank	40/50	1		06/25/16 22:58
166	T3.062516.230246	WG573484-03	Laboratory Control S	40/50	1		06/25/16 23:02
167	T3.062516.230624	WG573484-01	Reference Sample		1	L16060764-04	06/25/16 23:06
168	T3.062516.231017	WG573484-04	Matrix Spike	40/50	1	L16060764-04	06/25/16 23:10
169	T3.062516.231356	WG573484-05	Matrix Spike Duplica	40/50	1	L16060764-04	06/25/16 23:13
170	T3.062516.231733	L16060991-02	MW31-061616	40/50	10		06/25/16 23:17

Page: 5 Approved: June 27, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062516T3.1R.TXT
Analyst1: JYH Analyst2: N/A
Method: 200.7/6010B/6010C SOP: ME600G Rev: 8
Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838
ICSA: STD76382 ICSAB: STD76834 Int. Std: RGT31668
CCV: STD76511 LLCCV: STD76601 Tuning Sol :
Stannous : Hydroxylamine :

Workgroups: 573973,574055,574076,574216,574215,573952

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
171	T3.062516.232126	WG573952-03	Post Digestion Spike		10	L16060991-02	06/25/16 23:21
172	T3.062516.232503	WG573952-04	Serial Dilution		50	L16060991-02	06/25/16 23:25
173	T3.062516.232900	L16060995-01	UNTREATED	40/50	1000		06/25/16 23:29
174	T3.062516.233253	L16060995-02	AFTER	40/50	1000		06/25/16 23:32
175	T3.062516.233647	WG574249-43	CCV		1		06/25/16 23:36
176	T3.062516.234025	WG574249-44	CCB		1		06/25/16 23:40
177	T3.062516.234420	WG574249-45	Interference Check		1		06/25/16 23:44
178	T3.062516.234809	WG574249-46	Interference Check		1		06/25/16 23:48
179	T3.062516.235152	WG574249-47	CCV		1		06/25/16 23:51
180	T3.062516.235530	WG574249-48	CCB		1		06/25/16 23:55
181	T3.062516.235927	WG574249-49	Low Level Continuing Calibra		1		06/25/16 23:59

Comments

Seq.	Rerun	Dil.	Reason	Analytes
69				
Seq. 69-70: wrong WG numbers JYH				

Page: 6 Approved: June 27, 2016

Kim H. Rhoden

Microbac Laboratories Inc.

Data Checklist

Date: 23-JUN-2016
Analyst: JYH
Analyst: NA
Method: 6010B/6010C/200.7
Instrument: ICP-THERMO3
Curve Workgroup: 574093
Runlog ID: 75909
Analytical Workgroups: 573973,573749,573952

Add'l WG	
STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	X
ICB/CCB	X
ICSA/ICSAB	X
CRI	X
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	X
Client Forms	X
Level X	
Level 3	764,808
Level 4	1098,779,822,823,883,896,903,924
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	JYH
Secondary Reviewer	KHR
Comments	

Primary Reviewer:

Secondary Reviewer:
26-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-26-2016 07:48:33

Microbac Laboratories Inc.

Data Checklist

Date: 25-JUN-2016
Analyst: JYH
Analyst: NA
Method: 6010B/6010C/200.7
Instrument: ICP-THERMO3
Curve Workgroup: 574249
Runlog ID: 75923
Analytical Workgroups: 573973,574055,574076,574216,574215,573952

Add'l WG	
STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	X
ICB/CCB	X
ICSA/ICSAB	X
CRI	X
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	X
Client Forms	X
Level X	
Level 3	1178
Level 4	822,823,1148,1150,1151,1154,1155
	1227,1279,1289,986,787,991,995
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	JYH
Secondary Reviewer	KHR
Comments	

Primary Reviewer: Secondary Reviewer:
27-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-27-2016 14:55:21

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 6010C

AAB#: WG573973

Login Number: L16060823

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
126F-061416	02	06/14/16					06/17/2016	3.1	180		06/23/16	9.4	180	
126F-061416	02	06/14/16					06/17/2016	3.1	180		06/23/16	9.4	180	
126F-061416	02	06/14/16					06/17/2016	3.1	180		06/25/16	11.1	180	
C08-061416	03	06/14/16					06/17/2016	3	180		06/25/16	11	180	
C08-061416	03	06/14/16					06/17/2016	3	180		06/23/16	9.3	180	
C08-061416	03	06/14/16					06/17/2016	3	180		06/23/16	9.4	180	
C03-061416	04	06/14/16					06/17/2016	2.9	180		06/25/16	11	180	
C03-061416	04	06/14/16					06/17/2016	2.9	180		06/23/16	9.3	180	
C03-061416	04	06/14/16					06/17/2016	2.9	180		06/23/16	9.3	180	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4823910
Report generated 06/27/2016 10:56

METHOD BLANK SUMMARY

Login Number: L16060823
Blank File ID: T3.062316.192408
Prep Date: 06/17/16 12:25
Analyzed Date: 06/23/16 19:24
Analyst: JYH

Work Group: WG573973
Blank Sample ID: WG573190-03
Instrument ID: ICP-THERMO3
Method: 6010C

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573190-04	T3.062316.192807	06/23/16 19:28	01
126F-061416	L16060823-02	T3.062316.203318	06/23/16 20:33	01
C08-061416	L16060823-03	T3.062316.203719	06/23/16 20:37	01
C03-061416	L16060823-04	T3.062316.204123	06/23/16 20:41	01
126F-061416	L16060823-02	T3.062316.213215	06/23/16 21:32	DL01
C08-061416	L16060823-03	T3.062316.214348	06/23/16 21:43	DL01
C03-061416	L16060823-04	T3.062316.214744	06/23/16 21:47	DL01
LCS	WG573190-04	T3.062516.132214	06/25/16 13:22	02
126F-061416	L16060823-02	T3.062516.140359	06/25/16 14:03	02
C08-061416	L16060823-03	T3.062516.140759	06/25/16 14:07	02
C03-061416	L16060823-04	T3.062516.141200	06/25/16 14:12	02

Report Name: BLANK_SUMMARY
PDF File ID: 4823911
Report generated 06/27/2016 10:56

METHOD BLANK SUMMARY

Login Number: L16060823
Blank File ID: T3.062516.131816
Prep Date: 06/17/16 12:25
Analyzed Date: 06/25/16 13:18
Analyst: JYH

Work Group: WG573973
Blank Sample ID: WG573190-03
Instrument ID: ICP-THERMO3
Method: 6010C

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573190-04	T3.062316.192807	06/23/16 19:28	01
126F-061416	L16060823-02	T3.062316.203318	06/23/16 20:33	01
C08-061416	L16060823-03	T3.062316.203719	06/23/16 20:37	01
C03-061416	L16060823-04	T3.062316.204123	06/23/16 20:41	01
126F-061416	L16060823-02	T3.062316.213215	06/23/16 21:32	DL01
C08-061416	L16060823-03	T3.062316.214348	06/23/16 21:43	DL01
C03-061416	L16060823-04	T3.062316.214744	06/23/16 21:47	DL01
LCS	WG573190-04	T3.062516.132214	06/25/16 13:22	02
126F-061416	L16060823-02	T3.062516.140359	06/25/16 14:03	02
C08-061416	L16060823-03	T3.062516.140759	06/25/16 14:07	02
C03-061416	L16060823-04	T3.062516.141200	06/25/16 14:12	02

Report Name: BLANK_SUMMARY
PDF File ID: 4823911
Report generated 06/27/2016 10:56

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060823 Prep Date: 06/17/16 12:25 Sample ID: WG573190-03
Instrument ID: ICP-THERMO3 Run Date: 06/23/16 19:24 Prep Method: 3015
File ID: T3.062316.192408 Analyst: JYH Method: 6010C
Workgroup (AAB#): WG573973 Matrix: Water Units: mg/L
Contract #: Cal ID: ICP-TH-23-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Aluminum, Total	0.0500	0.200	0.0500	1	U
Beryllium, Total	0.00500	0.0200	0.00500	1	U
Calcium, Total	0.125	0.500	0.125	1	U
Iron, Total	0.0500	0.200	0.0500	1	U
Magnesium, Total	0.250	1.00	0.250	1	U
Potassium, Total	0.500	2.00	0.500	1	U
Sodium, Total	0.250	1.00	0.412	1	J

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name:BLANK
PDF ID: 4823912
07-JUL-2016 17:47

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060823 Prep Date: 06/17/16 12:25 Sample ID: WG573190-03
Instrument ID: ICP-THERMO3 Run Date: 06/25/16 13:18 Prep Method: 3015
File ID: T3.062516.131816 Analyst: JYH Method: 6010C
Workgroup (AAB#): WG573973 Matrix: Water Units: mg/L
Contract #: Cal ID: ICP-TH-25-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Selenium, Total	0.00500	0.0200	0.00784	1	J

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4823912
07-JUL-2016 17:47

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060823 Run Date: 06/23/2016 Sample ID: WG573190-04
Instrument ID: ICP-THERMO3 Run Time: 19:28 Prep Method: 3015
File ID: T3.062316.192807 Analyst: JYH Method: 6010C
Workgroup (AAB#): WG573973 Matrix: Water Units: mg/L
QC Key: DOD4 Lot#: STD76576 Cal ID: ICP-TH-23-JUN-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Aluminum, Total	6.25	6.24	99.9	80 - 120	
Beryllium, Total	0.0313	0.0311	99.6	80 - 120	
Calcium, Total	6.25	6.64	106	80 - 120	
Iron, Total	2.50	2.60	104	80 - 120	
Magnesium, Total	6.25	6.49	104	80 - 120	
Potassium, Total	31.3	32.2	103	80 - 120	
Sodium, Total	31.3	32.4	104	80 - 120	

LCS - Modified 03/06/2008
PDF File ID: 4823913
Report generated: 07/07/2016 17:47

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060823 Run Date: 06/25/2016 Sample ID: WG573190-04
Instrument ID: ICP-THERMO3 Run Time: 13:22 Prep Method: 3015
File ID: T3.062516.132214 Analyst: JYH Method: 6010C
Workgroup (AAB#): WG573973 Matrix: Water Units: mg/L
QC Key: DOD4 Lot#: STD76576 Cal ID: ICP-TH-25-JUN-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Selenium, Total	0.250	0.233	93.3	80 - 120	

LCS - Modified 03/06/2008
PDF File ID: 4823913
Report generated: 07/07/2016 17:47

Microbac Laboratories Inc.
MATRIX SPIKE AND MATRIX SPIKE DUP (MS/MSD)

Loginnum: L16060823 Cal ID: ICP-THERMO3 - Worknum: WG573973
Instrument ID: ICP-THERMO3 Contract #: Method: 6010C
Parent ID: WG573190-02 File ID: T3.062316.201412 Dil: 1 Matrix: WATER
Sample ID: WG573190-07 MS File ID: T3.062316.201808 Dil: 1 Units: mg/L
Sample ID: WG573190-08 MSD File ID: T3.062316.202147 Dil: 1

Analyte	Parent	MS Spiked	MS Found	MS %Rec	MSD Spiked	MSD Found	MSD %Rec	%RPD	%Rec Limits	RPD Limit	Q
Aluminum	ND	6.25	6.24	99.9	6.25	6.28	101	0.589	80 - 120	20	
Beryllium	ND	0.0313	0.0310	99.3	0.0313	0.0314	100	1.16	80 - 120	20	
Calcium	66.5	6.25	71.3	76.5	6.25	73.1	106	2.53	80 - 120	20	*
Iron	0.336	2.50	2.91	103	2.50	2.93	104	0.788	80 - 120	20	
Magnesium	18.8	6.25	25.0	99.9	6.25	25.4	106	1.47	80 - 120	20	
Potassium	1.76	31.3	33.3	101	31.3	33.9	103	1.61	80 - 120	20	
Sodium	19.9	31.3	51.5	101	31.3	52.3	104	1.67	80 - 120	20	

* FAILS %REC LIMIT

FAILS RPD LIMIT

NOTE: This is an internal quality control sample.

WG_MS_MSD_DRYWT - Modified 05/26/2011
PDF File ID: 4823914
Report generated 07/07/2016 17:47

Microbac Laboratories Inc.
MATRIX SPIKE AND MATRIX SPIKE DUP (MS/MSD)

Loginnum: L16060823 Cal ID: ICP-THERMO3 - Worknum: WG573973
Instrument ID: ICP-THERMO3 Contract #: Method: 6010C
Parent ID: WG573190-02 File ID: T3.062516.132554 Dil: 1 Matrix: WATER
Sample ID: WG573190-07 MS File ID: T3.062516.132950 Dil: 1 Units: mg/L
Sample ID: WG573190-08 MSD File ID: T3.062516.133327 Dil: 1

Analyte	Parent	MS Spiked	MS Found	MS %Rec	MSD Spiked	MSD Found	MSD %Rec	%RPD	%Rec Limits	RPD Limit	Q
Selenium	ND	0.250	0.241	96.4	0.250	0.240	96.0	0.437	80 - 120	20	

* FAILS %REC LIMIT

\# FAILS RPD LIMIT

NOTE: This is an internal quality control sample.

WG_MS_MSD_DRYWT - Modified 05/26/2011
PDF File ID: 4823914
Report generated 07/07/2016 17:47

Microbac Laboratories Inc.
Serial Dilution Report

Login: L16060823 **Worknum:** WG573973
Instrument: ICP-THERMO3 **Method:** 6010C
Serial Dil: WG573973-04 **File ID:** T3.062516.134830 **Dil:** 5 **Units:** ug/L
Sample: L16060822-02 **File ID:** T3.062516.134058 **Dil:** 1

Analyte	Sample	Qual	Serial Dil	Qual	% Diff	Q
Aluminum	7.00		ND	U		
Beryllium	0.100		0.600		500.00	E
Calcium	6220		5990		3.69	
Iron	2080		2120		1.93	
Magnesium	4400		4250		3.40	
Potassium	1500		1580		5.21	
Selenium	4.52		ND	U		
Sodium	83200		84000		0.89	

U = Result is below MDL.
F = Result is greater than or equal to MDL and less than the RL.
X = Result is greater than or equal to RL and less than 25 times the MDL.
E = %D exceeds control limit of 10% and initial sample result is greater than or equal to 25 times the MDL.

SERIAL_DIL - Modified 09/22/2008
PDF File ID: 4823908
06/27/2016 15:01

Microbac Laboratories Inc.
Serial Dilution Report

Login: L16060823 **Worknum:** WG573973
Instrument: ICP-THERMO3 **Method:** 6010C
Serial Dil: WG573973-02 **File ID:** T3.062316.195438 **Dil:** 5 **Units:** ug/L
Sample: L16060779-11 **File ID:** T3.062316.194701 **Dil:** 1

Analyte	Sample	Qual	Serial Dil	Qual	% Diff	Q
Aluminum	173		160		7.72	
Beryllium	0.0300	X	0.100	X	233.00	
Calcium	4280		3960		7.51	
Iron	98.8		139		40.50	E
Magnesium	726		289		60.30	E
Potassium	444		612		37.80	E
Selenium	ND	U	ND	U		
Sodium	142000		130000		8.61	

U = Result is below MDL.
F = Result is greater than or equal to MDL and less than the RL.
X = Result is greater than or equal to RL and less than 25 times the MDL.
E = %D exceeds control limit of 10% and initial sample result is greater than or equal to 25 times the MDL.

SERIAL_DIL - Modified 09/22/2008
PDF File ID: 4823908
06/27/2016 15:01

Microbac Laboratories Inc.
POST SPIKE REPORT

Sample Login ID: L16060823 **Worknum:** WG573973
Instrument ID: ICP-THERMO3 **Method:** 6010C
Post Spike ID: WG573973-01 **File ID:** T3.062316.195059 **Dil:** 1 **Units:** ug/L
Sample ID: L16060779-11 **File ID:** T3.062316.194701 **Dil:** 1 **Matrix:** Water

Analyte	Post Spike Result	C	Sample Result	C	Spike Added(SA)	% R	Control Limit %R	Q
ALUMINUM	5140		173		5000	99.6	75 - 125	
BERYLLIUM	25.0		0	U	25	99.8	75 - 125	
CALCIUM	8980		4280		5000	102.5	75 - 125	
IRON	2120		98.8	F	2000	101.7	75 - 125	
MAGNESIUM	5660		726	F	5000	100.1	75 - 125	
POTASSIUM	25600		444	F	25000	100.7	75 - 125	
SELENIUM	192		0	U	200	96.2	75 - 125	
SODIUM	153000		142000		25000	101.9	75 - 125	

N = % Recovery exceeds control limits
F = Result is between MDL and RL
U = Sample result is below MDL. A value of zero is used in the calculation

POST_SPIKE - Modified 03/06/2008
PDF File ID: 4823909
Report generated: 06/27/2016 15:01

Microbac Laboratories Inc.
POST SPIKE REPORT

Sample Login ID: L16060823 **Worknum:** WG573973
Instrument ID: ICP-THERMO3 **Method:** 6010C
Post Spike ID: WG573973-03 **File ID:** T3.062516.134452 **Dil:** 1 **Units:** ug/L
Sample ID: L16060822-02 **File ID:** T3.062516.134058 **Dil:** 1 **Matrix:** Water

Analyte	Post Spike Result	C	Sample Result	C	Spike Added(SA)	% R	Control Limit %R	Q
ALUMINUM	4880		0	U	5000	97.5	75 - 125	
BERYLLIUM	24.3		0	U	25	97.1	75 - 125	
CALCIUM	10600		6220		5000	101.0	75 - 125	
IRON	3850		2080		2000	99.0	75 - 125	
MAGNESIUM	8760		4400		5000	96.0	75 - 125	
POTASSIUM	25800		1500	F	25000	97.9	75 - 125	
SELENIUM	192		4.52	F	200	94.2	75 - 125	
SODIUM	100000		83200		25000	100.3	75 - 125	

N = % Recovery exceeds control limits
F = Result is between MDL and RL
U = Sample result is below MDL. A value of zero is used in the calculation

POST_SPIKE - Modified 03/06/2008
PDF File ID: 4823909
Report generated: 06/27/2016 15:01

Microbac Laboratories Inc.
Initial Calibration Summary

Login:	L16060823	Workgroup (AAB#):	WG573973
Analytical Method:	6010C	Instrument ID:	ICP-THERMO3
ICAL Worknum:	WG574093	Initial Calibration Date:	23-JUN-2016 16:45

	WG574093-01		WG574093-02		WG574093-03		WG574093-04		WG574093-05			
	Conc	INT	Conc	INT	Conc	INT	Conc	INT	Conc	INT	R	Q
ALUMINUM	0	0.00129	.1	0.00207	.2	0.00279	10	0.0950	20	0.188	.999993	
BERYLLIUM	0	0.000930	.0005	0.00138	.001	0.00179	.05	0.0552	.1	0.108	.99998	
CALCIUM	0	0.000280	.1	0.00433	.2	0.00735	10	0.497	20	0.993	.999968	
IRON	0	-0.000490	.04	0.000710	.08	0.000970	4	0.0998	8	0.200	.999657	
MAGNESIUM	0	-0.000700	NA	NA	.2	0.000790	10	0.0960	20	0.193	.999996	
POTASSIUM	0	-0.0106	.5	0.0183	1	0.0487	50	3.45	100	6.93	.999984	
SELENIUM	0	-0.000250	NA	NA	.008	-0.000140	.4	0.00687	.8	0.0139	.999993	
SODIUM	0	-0.0645	.5	0.0202	1	0.0962	50	9.63	100	19.4	.999985	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
*** = Out of Compliance; R < 0.995**

INT_CAL_ICP - Modified 03/06/2008
PDF File ID: 4823917
Report generated: 27-JUN-2016 15:01

Microbac Laboratories Inc.
Initial Calibration Summary

Login:	L16060823	Workgroup (AAB#):	WG573973
Analytical Method:	6010C	Instrument ID:	ICP-THERMO3
ICAL Worknum:	WG574249	Initial Calibration Date:	25-JUN-2016 12:36

	WG574249-01		WG574249-02		WG574249-03		WG574249-04		WG574249-05			
	Conc	INT	Conc	INT	Conc	INT	Conc	INT	Conc	INT	R	Q
ALUMINUM	0	0.000800	.1	0.00148	.2	0.00220	10	0.0828	20	0.162	.99995	
BERYLLIUM	0	0.000480	.0005	0.000970	.001	0.00139	.05	0.0482	.1	0.0948	.999766	
CALCIUM	0	0.00288	.1	0.00582	.2	0.00931	10	0.466	20	0.917	.99989	
IRON	0	-0.000710	.04	0.000330	.08	0.00148	4	0.0965	8	0.191	.99933	
MAGNESIUM	0	-0.000260	NA	NA	.2	0.00109	10	0.0796	20	0.159	.99999	
POTASSIUM	0	-0.00661	.5	0.0195	1	0.0419	50	3.07	100	6.10	.999988	
SELENIUM	0	-0.000510	NA	NA	.008	-0.000340	.4	0.00649	.8	0.0134	.999238	
SODIUM	0	-0.0478	.5	0.0267	1	0.108	50	9.20	100	18.4	.99999	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
*** = Out of Compliance; R < 0.995**

INT_CAL_ICP - Modified 03/06/2008
PDF File ID: 4823917
Report generated: 27-JUN-2016 15:01

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-07
Instrument ID: ICP-THERMO3
Run Time: 16:52
Method: 6010C
File ID: T3.062316.165229
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-THERI - 23-JUN-16
Matrix: WATER

Analytes	MDL	RDL	Concentration	Qualifier
ALUMINUM	.04	.16	.04	U
BERYLLIUM	.004	.016	.004	U
CALCIUM	.1	.4	.1	U
IRON	.04	.16	.04	U
MAGNESIUM	.2	.8	.2	U
POTASSIUM	.4	1.6	.4	U
SELENIUM	.004	.016	-.00449	F
SODIUM	.2	.8	.2	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4823919
Report generated 06/27/2016 15:02

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16060823
Instrument ID: ICP-THERMO3
File ID: T3.062516.124348
Workgroup (AAB#): WG573973
Matrix: WATER

Run Date: 06/25/2016
Run Time: 12:43
Analyst: JYH
Cal ID: ICP-THERI - 25-JUN-16

Sample ID: WG574249-07
Method: 6010C
Units: mg/L

Analytes	MDL	RDL	Concentration	Qualifier
ALUMINUM	.04	.16	.04	U
BERYLLIUM	.004	.016	.004	U
CALCIUM	.1	.4	.1	U
IRON	.04	.16	.04	U
MAGNESIUM	.2	.8	.2	U
POTASSIUM	.4	1.6	.4	U
SELENIUM	.004	.016	.0087	*
SODIUM	.2	.8	.2	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4823919
Report generated 06/27/2016 15:02

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-13
Instrument ID: ICP-THERMO3
Run Time: 17:15
Method: 6010C
File ID: T3.062316.171523
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-19
Instrument ID: ICP-THERMO3
Run Time: 19:20
Method: 6010C
File ID: T3.062316.192008
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.423	F

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-21
Instrument ID: ICP-THERMO3
Run Time: 20:06
Method: 6010C
File ID: T3.062316.200616
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-23
Instrument ID: ICP-THERMO3
Run Time: 20:52
Method: 6010C
File ID: T3.062316.205247
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	-0.00415	F
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-25
Instrument ID: ICP-THERMO3
Run Time: 21:39
Method: 6010C
File ID: T3.062316.213949
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-27
Instrument ID: ICP-THERMO3
Run Time: 22:23
Method: 6010C
File ID: T3.062316.222310
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	-0.00859	F
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-13
Instrument ID: ICP-THERMO3
Run Time: 13:06
Method: 6010C
File ID: T3.062516.130641
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00482	F
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-15
Instrument ID: ICP-THERMO3
Run Time: 14:00
Method: 6010C
File ID: T3.062516.140001
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-17
Instrument ID: ICP-THERMO3
Run Time: 14:46
Method: 6010C
File ID: T3.062516.144639
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00512	F
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-44
Instrument ID: ICP-THERMO3
Run Time: 23:40
Method: 6010C
File ID: T3.062516.234025
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-48
Instrument ID: ICP-THERMO3
Run Time: 23:55
Method: 6010C
File ID: T3.062516.235530
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4823922
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-06
Instrument ID: ICP-THERMO3
Run Time: 16:48
Method: 6010C
File ID: T3.062316.164851
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Aluminum		10	10.0	100	90 - 110	
Beryllium		.05	0.0505	101	90 - 110	
Calcium		10	10.0	100	90 - 110	
Iron		4	4.01	100	90 - 110	
Magnesium		10	10.1	101	90 - 110	
Potassium		50	50.2	100	90 - 110	
Selenium		.4	0.399	99.8	90 - 110	
Sodium		50	50.1	100	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4823918
Report generated 06/27/2016 15:01

Microbac Laboratories Inc.

INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16060823 Run Date: 06/25/2016 Sample ID: WG574249-06
Instrument ID: ICP-THERMO3 Run Time: 12:40 Method: 6010C
File ID: T3.062516.124009 Analyst: JYH Units: mg/L
Workgroup (AAB#): WG573973 Cal ID: ICP-TH - 25-JUN-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Aluminum		10	10.1	101	90 - 110	
Beryllium		.05	0.0504	101	90 - 110	
Calcium		10	10.1	101	90 - 110	
Iron		4	4.05	101	90 - 110	
Magnesium		10	10.1	101	90 - 110	
Potassium		50	50.3	101	90 - 110	
Selenium		.4	0.400	100	90 - 110	
Sodium		50	50.1	100	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4823918
Report generated 06/27/2016 15:01

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-12
Instrument ID: ICP-THERMO3
Run Time: 17:11
Method: 6010C
File ID: T3.062316.171145
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	10.1	mg/L	101	90 - 110		
Beryllium		0.0500	0.0502	mg/L	100	90 - 110		
Calcium		10.0	10.1	mg/L	101	90 - 110		
Iron		4.00	4.06	mg/L	102	90 - 110		
Magnesium		10.0	10.2	mg/L	102	90 - 110		
Potassium		50.0	50.0	mg/L	100	90 - 110		
Selenium		0.400	0.397	mg/L	99.3	90 - 110		
Sodium		50.0	49.9	mg/L	99.8	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-18
Instrument ID: ICP-THERMO3
Run Time: 19:16
Method: 6010C
File ID: T3.062316.191630
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	9.45	mg/L	94.5	90 - 110		
Beryllium		0.0500	0.0476	mg/L	95.2	90 - 110		
Calcium		10.0	9.56	mg/L	95.6	90 - 110		
Iron		4.00	3.80	mg/L	95.0	90 - 110		
Magnesium		10.0	9.52	mg/L	95.2	90 - 110		
Potassium		50.0	47.8	mg/L	95.7	90 - 110		
Selenium		0.400	0.378	mg/L	94.5	90 - 110		
Sodium		50.0	48.1	mg/L	96.3	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-20
Instrument ID: ICP-THERMO3
Run Time: 20:02
Method: 6010C
File ID: T3.062316.200238
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	9.57	mg/L	95.7	90 - 110		
Beryllium		0.0500	0.0483	mg/L	96.5	90 - 110		
Calcium		10.0	9.59	mg/L	95.9	90 - 110		
Iron		4.00	3.86	mg/L	96.5	90 - 110		
Magnesium		10.0	9.65	mg/L	96.5	90 - 110		
Potassium		50.0	48.3	mg/L	96.5	90 - 110		
Selenium		0.400	0.384	mg/L	95.9	90 - 110		
Sodium		50.0	47.9	mg/L	95.9	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-22
Instrument ID: ICP-THERMO3
Run Time: 20:49
Method: 6010C
File ID: T3.062316.204910
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	9.91	mg/L	99.1	90 - 110		
Beryllium		0.0500	0.0500	mg/L	99.9	90 - 110		
Calcium		10.0	9.85	mg/L	98.5	90 - 110		
Iron		4.00	3.92	mg/L	98.0	90 - 110		
Magnesium		10.0	9.86	mg/L	98.6	90 - 110		
Potassium		50.0	49.3	mg/L	98.6	90 - 110		
Selenium		0.400	0.405	mg/L	101	90 - 110		
Sodium		50.0	49.2	mg/L	98.4	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-24
Instrument ID: ICP-THERMO3
Run Time: 21:36
Method: 6010C
File ID: T3.062316.213611
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	9.40	mg/L	94.0	90 - 110		
Beryllium		0.0500	0.0475	mg/L	95.0	90 - 110		
Calcium		10.0	9.50	mg/L	95.0	90 - 110		
Iron		4.00	3.81	mg/L	95.4	90 - 110		
Magnesium		10.0	9.53	mg/L	95.3	90 - 110		
Potassium		50.0	47.4	mg/L	94.9	90 - 110		
Selenium		0.400	0.380	mg/L	95.0	90 - 110		
Sodium		50.0	47.3	mg/L	94.6	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-26
Instrument ID: ICP-THERMO3
Run Time: 22:19
Method: 6010C
File ID: T3.062316.221931
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	9.29	mg/L	92.9	90 - 110		
Beryllium		0.0500	0.0467	mg/L	93.4	90 - 110		
Calcium		10.0	9.32	mg/L	93.2	90 - 110		
Iron		4.00	3.74	mg/L	93.4	90 - 110		
Magnesium		10.0	9.28	mg/L	92.8	90 - 110		
Potassium		50.0	46.9	mg/L	93.8	90 - 110		
Selenium		0.400	0.374	mg/L	93.5	90 - 110		
Sodium		50.0	46.6	mg/L	93.1	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-12
Instrument ID: ICP-THERMO3
Run Time: 13:03
Method: 6010C
File ID: T3.062516.130304
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	10.0	mg/L	100	90 - 110		
Beryllium		0.0500	0.0498	mg/L	99.6	90 - 110		
Calcium		10.0	9.94	mg/L	99.4	90 - 110		
Iron		4.00	4.00	mg/L	99.9	90 - 110		
Magnesium		10.0	9.92	mg/L	99.2	90 - 110		
Potassium		50.0	49.2	mg/L	98.4	90 - 110		
Selenium		0.400	0.405	mg/L	101	90 - 110		
Sodium		50.0	49.3	mg/L	98.6	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-14
Instrument ID: ICP-THERMO3
Run Time: 13:56
Method: 6010C
File ID: T3.062516.135622
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	9.88	mg/L	98.8	90 - 110		
Beryllium		0.0500	0.0488	mg/L	97.5	90 - 110		
Calcium		10.0	9.94	mg/L	99.4	90 - 110		
Iron		4.00	3.96	mg/L	98.9	90 - 110		
Magnesium		10.0	9.70	mg/L	97.0	90 - 110		
Potassium		50.0	49.1	mg/L	98.2	90 - 110		
Selenium		0.400	0.395	mg/L	98.8	90 - 110		
Sodium		50.0	49.4	mg/L	98.8	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-16
Instrument ID: ICP-THERMO3
Run Time: 14:43
Method: 6010C
File ID: T3.062516.144313
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	10.5	mg/L	105	90 - 110		
Beryllium		0.0500	0.0519	mg/L	104	90 - 110		
Calcium		10.0	10.5	mg/L	105	90 - 110		
Iron		4.00	4.13	mg/L	103	90 - 110		
Magnesium		10.0	10.2	mg/L	102	90 - 110		
Potassium		50.0	50.7	mg/L	101	90 - 110		
Selenium		0.400	0.413	mg/L	103	90 - 110		
Sodium		50.0	51.4	mg/L	103	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-43
Instrument ID: ICP-THERMO3
Run Time: 23:36
Method: 6010C
File ID: T3.062516.233647
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	10.2	mg/L	102	90 - 110		
Beryllium		0.0500	0.0497	mg/L	99.5	90 - 110		
Calcium		10.0	10.4	mg/L	104	90 - 110		
Iron		4.00	4.06	mg/L	102	90 - 110		
Magnesium		10.0	9.72	mg/L	97.2	90 - 110		
Potassium		50.0	51.1	mg/L	102	90 - 110		
Selenium		0.400	0.405	mg/L	101	90 - 110		
Sodium		50.0	52.1	mg/L	104	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-47
Instrument ID: ICP-THERMO3
Run Time: 23:51
Method: 6010C
File ID: T3.062516.235152
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	10.3	mg/L	103	90 - 110		
Beryllium		0.0500	0.0502	mg/L	100	90 - 110		
Calcium		10.0	10.5	mg/L	105	90 - 110		
Iron		4.00	4.08	mg/L	102	90 - 110		
Magnesium		10.0	9.81	mg/L	98.1	90 - 110		
Potassium		50.0	51.1	mg/L	102	90 - 110		
Selenium		0.400	0.412	mg/L	103	90 - 110		
Sodium		50.0	51.8	mg/L	104	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4823921
Report generated 06/27/2016 15:03

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-08
Instrument ID: ICP-THERMO3
Run Time: 16:56
Method: 6010C
File ID: T3.062316.165617
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		0.160	0.162	mg/L	101	70 - 130		
Beryllium		0.00160	0.00155	mg/L	96.9	70 - 130		
Calcium		0.400	0.451	mg/L	113	70 - 130		
Iron		0.0800	0.0758	mg/L	94.7	70 - 130		
Magnesium		0.400	0.430	mg/L	108	70 - 130		
Potassium		0.800	0.815	mg/L	102	70 - 130		
Sodium		0.400	0.447	mg/L	112	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4823923
Report generated 06/27/2016 15:04

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060823
Run Date: 06/23/2016
Sample ID: WG574093-28
Instrument ID: ICP-THERMO3
Run Time: 22:27
Method: 6010C
File ID: T3.062316.222708
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 23-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		0.160	0.137	mg/L	85.5	70 - 130		
Beryllium		0.00160	0.00136	mg/L	85.0	70 - 130		
Calcium		0.400	0.387	mg/L	96.8	70 - 130		
Iron		0.0800	0.0604	mg/L	75.5	70 - 130		
Magnesium		0.400	0.355	mg/L	88.8	70 - 130		
Potassium		0.800	0.745	mg/L	93.1	70 - 130		
Sodium		0.400	0.436	mg/L	109	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4823923
Report generated 06/27/2016 15:04

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-08
Instrument ID: ICP-THERMO3
Run Time: 12:47
Method: 6010C
File ID: T3.062516.124745
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		0.160	0.165	mg/L	103	70 - 130		
Beryllium		0.00160	0.00166	mg/L	104	70 - 130		
Calcium		0.400	0.414	mg/L	103	70 - 130		
Iron		0.0800	0.0939	mg/L	117	70 - 130		
Magnesium		0.400	0.441	mg/L	110	70 - 130		
Potassium		0.800	0.887	mg/L	111	70 - 130		
Selenium		0.0160	0.0204	mg/L	127	70 - 130		
Sodium		0.400	0.442	mg/L	110	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4823923
Report generated 06/27/2016 15:04

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060823
Run Date: 06/25/2016
Sample ID: WG574249-26
Instrument ID: ICP-THERMO3
Run Time: 17:25
Method: 6010C
File ID: T3.062516.172540
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG573973
Cal ID: ICP-TH - 25-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		0.160	0.163	mg/L	102	70 - 130		
Beryllium		0.00160	0.00172	mg/L	108	70 - 130		
Calcium		0.400	0.375	mg/L	93.8	70 - 130		
Iron		0.0800	0.0950	mg/L	119	70 - 130		
Potassium		0.800	0.714	mg/L	89.2	70 - 130		
Selenium		0.0160	0.0209	mg/L	131	70 - 130		*
Sodium		0.400	0.408	mg/L	102	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4823923
Report generated 06/27/2016 15:04

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16060823
Instrument ID: ICP-THERMO3
Sol. A : WG574093-10 **File ID:** T3.062316.170408
Sol. AB : WG574093-11 **File ID:** T3.062316.170801

Workgroup (AAB#): WG573973
Method: 6010C
Units: mg/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Aluminum	250	266	106	250	265	106	
Beryllium	NS	0	NS	0.250	0.248	99.2	
Calcium	250	253	101	250	252	101	
Iron	100	99.7	99.7	100	97.4	97.4	
Magnesium	250	253	101	250	248	99.2	
Potassium	NS	0.163	NS	5.00	5.24	105	
Selenium	NS	-0.00219	NS	0.250	0.242	96.8	
Sodium	NS	0.0453	NS	5.00	5.37	107	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4823920
Report generated 06/27/2016 15:02

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16060823
Instrument ID: ICP-THERMO3
Sol. A : WG574249-10 **File ID:** T3.062516.125529
Sol. AB : WG574249-11 **File ID:** T3.062516.125921

Workgroup (AAB#): WG573973
Method: 6010C
Units: mg/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Aluminum	250	266	106	250	268	107	
Beryllium	NS	0.000110	NS	0.250	0.245	98.0	
Calcium	250	244	97.6	250	247	98.8	
Iron	100	96.8	96.8	100	96.5	96.5	
Magnesium	250	249	99.6	250	249	99.6	
Potassium	NS	0.0533	NS	5.00	5.19	104	
Selenium	NS	-0.00224	NS	0.250	0.249	99.6	
Sodium	NS	0.0404	NS	5.00	5.32	106	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4823920
Report generated 06/27/2016 15:02

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16060823
Instrument ID: ICP-THERMO3
Sol. A : WG574249-45 **File ID:** T3.062516.234420
Sol. AB : WG574249-46 **File ID:** T3.062516.234809

Workgroup (AAB#): WG573973
Method: 6010C
Units: mg/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Aluminum	250	263	105	250	271	108	
Beryllium	NS	0.000140	NS	0.250	0.244	97.6	
Calcium	250	248	99.2	250	254	102	
Iron	100	95.6	95.6	100	96.1	96.1	
Magnesium	250	236	94.4	250	239	95.6	
Potassium	NS	-0.0460	NS	5.00	5.21	104	
Selenium	NS	-0.00782	NS	0.250	0.238	95.2	
Sodium	NS	0.0818	NS	5.00	5.50	110	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4823920
Report generated 06/27/2016 15:02

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16060823 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	AG	AL	AS	B	BA
ALUMINUM	308.20	0	0	0	0	0
ANTIMONY	206.80	0	0.0000410	0	0	0
ARSENIC	189.00	0	0	0	0	0
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	0	0	0	0
BORON	249.60	0	0	0	0	0
CADMIUM	228.80	0	0	0.0115	0	-0.0000800
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0	0	0	0	0
COBALT	228.60	0	0	0	0	0
COPPER	224.70	0	0	0	0	0
IRON	261.10	0	0	0	0	0
LEAD	220.30	0	0.000260	0	0	0
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	0	0	0	0	0
MANGANESE	257.60	0	0	0	0	0
MOLYBDENUM	202.00	0	0	0	0	0
NICKEL	231.60	0	0	0	0	0
PHOSPHORUS	214.90	0	-0.000289	0	0	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0	-0.0000490	0	0	0
SILICON	212.40	0	0	0	0	0
SILVER	328.00	0	0	0	0	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0	0	0	0
THALLIUM	190.80	0	-0.0000120	0	0	0
TIN	189.90	0	0	0	0	0
TITANIUM	337.20	0	0	0	0	0
VANADIUM	292.40	0	0	0	0	0
ZINC	206.20	0	0.0000300	0	0	0
ZIRCONIUM	339.10	0	0	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4823916
Report generated: 06/27/2016 10:56

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16060823 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	BE	CA	CD	CO	CR
ALUMINUM	308.20	0	0	0	-0.000820	0
ANTIMONY	206.80	0	0	0	0	0.0200
ARSENIC	189.00	0	0	0	0	-0.00190
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	0	0	0	0
BORON	249.60	0	0	0	0.00343	0
CADMIUM	228.80	0	0	0	-0.00390	0
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0	0	0	0	0
COBALT	228.60	0	0	0	0	-0.000200
COPPER	224.70	0	0	0	0.0000770	-0.00100
IRON	261.10	0	0	0	0	-0.00100
LEAD	220.30	0	0	0	-0.0000130	-0.000132
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	0	0	0	0	0
MANGANESE	257.60	0	0	0	0	0.0000500
MOLYBDENUM	202.00	0	0	0	0	0
NICKEL	231.60	0	0	0	-0.000860	0
PHOSPHORUS	214.90	0	0	0	0	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0	0	0	0	0
SILICON	212.40	0	0	0	0	0
SILVER	328.00	0	0	0	0	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0.00000500	0	0	0
THALLIUM	190.80	0	0	0	0.00240	0.000276
TIN	189.90	0	0	0	0	0
TITANIUM	337.20	0	0	0	0	0
VANADIUM	292.40	0	0	0	0	-0.00480
ZINC	206.20	0	0	0	0	-0.00180
ZIRCONIUM	339.10	0	0	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4823916
Report generated: 06/27/2016 10:56

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16060823 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	CU	FE	K	LI	MG
ALUMINUM	308.20	0	0	0	0	0
ANTIMONY	206.80	0	0.0000560	0	0	0
ARSENIC	189.00	0	-0.0000500	0	0	0
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	0	0	0	0
BORON	249.60	0	0.000300	0	0	0
CADMIUM	228.80	0	-0.0000190	0	0	0
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0	0.0000500	0	0	0
COBALT	228.60	0	0	0	0	0
COPPER	224.70	0	0.00160	0	0	0
IRON	261.10	0	0	0	0	0
LEAD	220.30	0.000609	0	0	0	0
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	0	0	0	0	0
MANGANESE	257.60	0	0	0	0	0.00000300
MOLYBDENUM	202.00	0	0	0	0	0
NICKEL	231.60	0	0.0000420	0	0	0
PHOSPHORUS	214.90	-0.323	0.000900	0	0	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0	0	0	0	0
SILICON	212.40	0	0	0	0	0
SILVER	328.00	0	-0.000270	0	0	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0	0	0	0
THALLIUM	190.80	0	0	0	0	0
TIN	189.90	0	0	0	0	0
TITANIUM	337.20	0	-0.000400	0	0	0
VANADIUM	292.40	0	0.00000700	0	0	0
ZINC	206.20	0	0	0	0	0
ZIRCONIUM	339.10	0	-0.0000300	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4823916
Report generated: 06/27/2016 10:56

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16060823 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	MN	MO	NA	NI	P
ALUMINUM	308.20	0	0.0163	0	0	0
ANTIMONY	206.80	0	-0.00310	0	-0.00350	0
ARSENIC	189.00	0	0.00120	0	0	0
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	0	0	0	0
BORON	249.60	0	-0.00190	0	0	0
CADMIUM	228.80	0	0.0000320	0	-0.000770	0
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0.000360	0	0	0	0
COBALT	228.60	0	-0.00200	0	0.000100	0
COPPER	224.70	0	0.00160	0	-0.0123	0
IRON	261.10	0	0	0	0	0
LEAD	220.30	0	-0.00210	0	0.000110	0
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	-0.00290	-0.0230	0	0	0
MANGANESE	257.60	0	0.0000300	0	0	0
MOLYBDENUM	202.00	0	0	0	0	0
NICKEL	231.60	0	0	0	0	0
PHOSPHORUS	214.90	0	0.00710	0	0	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0.000600	0.000580	0	0	0
SILICON	212.40	0	0.0187	0	0	0
SILVER	328.00	0	-0.0000430	0	0	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0	0	0	0
THALLIUM	190.80	0.00100	0	0	0	0
TIN	189.90	0	0	0	0	0
TITANIUM	337.20	0	-0.000153	0	0	0
VANADIUM	292.40	-0.000200	-0.00830	0	0	0
ZINC	206.20	0	0	0	0	0
ZIRCONIUM	339.10	0	0	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4823916
Report generated: 06/27/2016 10:56

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16060823 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	PB	SB	SE	SI	SN
ALUMINUM	308.20	0	0	0	0	0
ANTIMONY	206.80	0	0	0	0	-0.0220
ARSENIC	189.00	0	0	0	0	0
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	0	0	0	0
BORON	249.60	0	0	0	0	0
CADMIUM	228.80	0	0	0	0	0
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0	0	0	0	0
COBALT	228.60	0	0	0	0	0
COPPER	224.70	0.00440	0	0	0	0
IRON	261.10	0	0	0	0	0
LEAD	220.30	0	0	0	0	0
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	0	0	0	0	0
MANGANESE	257.60	0	0	0	0	0
MOLYBDENUM	202.00	0	0	0	0	0
NICKEL	231.60	0	0	0	0	0
PHOSPHORUS	214.90	0	0	0	0	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0	0	0	0	0
SILICON	212.40	0	0	0	0	0
SILVER	328.00	0	0	0	0	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0	0	0	0
THALLIUM	190.80	0	0	0	0	0
TIN	189.90	0	0	0	0	0
TITANIUM	337.20	0	0	0	0	0
VANADIUM	292.40	0	0	0	0	0
ZINC	206.20	0	0	0	0	0
ZIRCONIUM	339.10	0	0	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4823916
Report generated: 06/27/2016 10:56

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16060823 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	SR	TI	TL	V	ZN
ALUMINUM	308.20	0	0	0	0.0950	0
ANTIMONY	206.80	0	0.00110	0	-0.00360	0
ARSENIC	189.00	0	0	0	0.000107	0
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	-0.00000700	0	0.000990	0
BORON	249.60	0	0	0	0	0
CADMIUM	228.80	0	0	0	0.000102	0
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0	0.0000550	0	0	0
COBALT	228.60	0	0.00210	0	0.0000200	0
COPPER	224.70	0	0.000269	0	0	0
IRON	261.10	0	0	0	0	0
LEAD	220.30	0	0	0	-0.000126	0
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	0	-0.00290	0	0	0
MANGANESE	257.60	0	0	0	0	0
MOLYBDENUM	202.00	0	0	0	-0.000110	0
NICKEL	231.60	0	0	0	0	0
PHOSPHORUS	214.90	0	0	0	-0.00100	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0	0	0	0	0
SILICON	212.40	0	0	0	0	0
SILVER	328.00	0	-0.000720	0	-0.000260	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0	0	0	0
THALLIUM	190.80	0	-0.000800	0	-0.00490	0
TIN	189.90	0	-0.00190	0	0	0
TITANIUM	337.20	0	0	0	0	0
VANADIUM	292.40	0	0.000820	0	0	0
ZINC	206.20	0	0	0	0	0
ZIRCONIUM	339.10	0	0	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4823916
Report generated: 06/27/2016 10:56

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16060823 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	ZR
ALUMINUM	308.20	0
ANTIMONY	206.80	0
ARSENIC	189.00	0
BARIUM	455.40	0
BERYLLIUM	313.10	0
BORON	249.60	0
CADMIUM	228.80	0
CALCIUM	422.60	0
CHROMIUM	267.70	0
COBALT	228.60	0
COPPER	224.70	0
IRON	261.10	0
LEAD	220.30	0
LITHIUM	670.70	0
MAGNESIUM	279.00	0
MANGANESE	257.60	0
MOLYBDENUM	202.00	0
NICKEL	231.60	0
PHOSPHORUS	214.90	0
POTASSIUM	766.40	0
SELENIUM	196.00	0
SILICON	212.40	0
SILVER	328.00	0
SODIUM	589.50	0
STRONTIUM	407.70	0
THALLIUM	190.80	0
TIN	189.90	0
TITANIUM	337.20	0
VANADIUM	292.40	0
ZINC	206.20	0
ZIRCONIUM	339.10	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4823916
Report generated: 06/27/2016 10:56

Microbac Laboratories Inc.
LINEAR RANGE (QUARTERLY)

Login Number: L16060823 **Date:** 04/29/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Integration Time (Sec.)	Concentration (mg/L)
Aluminum	10.00	900.0
Antimony	20.00	45.0
Arsenic	10.00	45.0
Barium	10.00	45.0
Beryllium	10.00	1.8
Boron	20.00	45.0
Cadmium	20.00	4.5
Calcium	5.00	270.0
Chromium	20.00	36.0
Cobalt	20.00	45.0
Copper	20.00	180.0
Iron	5.00	720.0
Lead	20.00	225.0
Lithium	5.00	36.0
Magnesium	5.00	900.0
Manganese	10.00	36.0
Molybdenum	20.00	27.0
Nickel	20.00	90.0
Phosphorus	20.00	180.0
Potassium	5.00	450.0
Selenium	20.00	90.0
Silicon	20.00	36.0
Silver	10.00	9.0
Sodium	5.00	270.0
Strontium	5.00	9.0
Thallium	20.00	18.0
Tin	20.00	45.0
Titanium	5.00	36.0
Vanadium	20.00	27.0
Zinc	20.00	45.0
Zirconium	10.00	45.0

Comments:
All analytes passed acceptance criteria at the specified concentration.

LINEAR_RANGE - Modified 03/06/2008
PDF File ID: 4823915
Report generated: 06/27/2016 10:56

2.3.1.3 Raw Data

Element, Wavelength and Order	Date of Fit	Date of Cal.	Type of Fit	Weighting	A0	A1	A2	n (Exponent)
Ag 328.068 {103}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000292	0.057216	0.000000	1.000000
Al 308.215 {109}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.001288	0.009253	0.000000	1.000000
As 189.042 {478}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000077	0.027152	0.000000	1.000000
B 249.678 {135}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.000040	0.025100	0.000000	1.000000
Ba 455.403 { 74}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.031526	2.349165	0.000000	1.000000
Be 313.107 {108}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.000934	1.055405	0.000000	1.000000
Ca 422.673 { 80}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.000283	0.049609	0.000000	1.000000
Cd 228.802 {447}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000142	0.651471	0.000000	1.000000
Co 228.616 {447}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000247	0.409702	0.000000	1.000000
Cr 267.716 {126}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.000367	0.066578	0.000000	1.000000
Cu 224.700 {450}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000343	0.142601	0.000000	1.000000
Fe 261.187 {129}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000490	0.025122	0.000000	1.000000
K 766.490 { 44}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.010609	0.069349	0.000000	1.000000
Li 670.784 { 50}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.011827	1.674850	0.000000	1.000000
Mg 279.079 {121}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000702	0.009699	0.000000	1.000000
Mn 257.610 {131}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000277	0.317266	0.000000	1.000000
Mo 202.030 {467}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.000054	0.203043	0.000000	1.000000
Na 589.592 { 57}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.064533	0.194326	0.000000	1.000000
Ni 231.604 {446}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000942	0.155561	0.000000	1.000000
P 214.914 {457}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000263	0.015029	0.000000	1.000000
Pb 220.353 {453}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000678	0.070221	0.000000	1.000000
Sb 206.833 {463}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.000010	0.034653	0.000000	1.000000
Se 196.090 {472}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000253	0.017694	0.000000	1.000000
Si 212.412 {459}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.002512	0.043862	0.000000	1.000000
Sn 189.989 {477}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.000061	0.078522	0.000000	1.000000
Sr 407.771 { 83}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.002450	3.941080	0.000000	1.000000
Ti 337.280 {100}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.003027	0.134433	0.000000	1.000000
Tl 190.856 {477}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.000414	0.032326	0.000000	1.000000
V 292.402 {115}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.000001	0.079042	0.000000	1.000000
Y 224.306 {450}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Y 360.073 { 94}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Y 377.433 { 89}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Zn 206.200 {463}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	0.000166	0.468267	0.000000	1.000000
Zr 339.198 { 99}	6/23/2016 16:48:46	6/23/2016 16:48:46	Linear	1/Conc	-0.019987	0.006947	0.000000	1.000000

Element, Wavelength and Order	Correlation	Std Error of Est	Predicted MDL	Predicted MQL	Status	Reslope		QC Norm	
						Slope	Y-int	Slope factor	Offset
Ag 328.068 {103}	0.999400	0.000005	0.002814	0.009379	OK.	1.000000	0.000000	1	0
Al 308.215 {109}	0.999993	0.000002	0.011897	0.039658	OK.	1.000000	0.000000	1	0
As 189.042 {478}	0.999595	0.000003	0.004016	0.013387	OK.	1.000000	0.000000	1	0
B 249.678 {135}	0.999985	0.000001	0.003782	0.012607	OK.	1.000000	0.000000	1	0
Ba 455.403 { 74}	0.999987	0.000075	0.001344	0.004481	OK.	1.000000	0.000000	1	0
Be 313.107 {108}	0.999980	0.000002	0.000124	0.000413	OK.	1.000000	0.000000	1	0
Ca 422.673 { 80}	0.999968	0.000025	0.053022	0.176739	OK.	1.000000	0.000000	1	0
Cd 228.802 {447}	0.999778	0.000005	0.000338	0.001128	OK.	1.000000	0.000000	1	0
Co 228.616 {447}	0.999932	0.000006	0.000516	0.001721	OK.	1.000000	0.000000	1	0
Cr 267.716 {126}	0.999922	0.000003	0.001537	0.005123	OK.	1.000000	0.000000	1	0
Cu 224.700 {450}	0.999931	0.000005	0.001747	0.005824	OK.	1.000000	0.000000	1	0
Fe 261.187 {129}	0.999657	0.000017	0.028560	0.095199	OK.	1.000000	0.000000	1	0
K 766.490 { 44}	0.999984	0.000125	0.120772	0.402574	OK.	1.000000	0.000000	1	0
Li 670.784 { 50}	0.999884	0.000250	0.004948	0.016494	OK.	1.000000	0.000000	1	0
Mg 279.079 {121}	0.999996	0.000003	0.098587	0.328625	OK.	1.000000	0.000000	1	0
Mn 257.610 {131}	0.999409	0.000035	0.003174	0.010580	OK.	1.000000	0.000000	1	0
Mo 202.030 {467}	0.999985	0.000007	0.000553	0.001844	OK.	1.000000	0.000000	1	0
Na 589.592 { 57}	0.999985	0.000335	0.039834	0.132780	OK.	1.000000	0.000000	1	0
Ni 231.604 {446}	0.999721	0.000012	0.001443	0.004811	OK.	1.000000	0.000000	1	0
P 214.914 {457}	0.999997	0.000002	0.009293	0.030977	OK.	1.000000	0.000000	1	0
Pb 220.353 {453}	0.999947	0.000002	0.004623	0.015411	OK.	1.000000	0.000000	1	0
Sb 206.833 {463}	0.999906	0.000004	0.005569	0.018563	OK.	1.000000	0.000000	1	0
Se 196.090 {472}	0.999993	0.000000	0.008923	0.029743	OK.	1.000000	0.000000	1	0
Si 212.412 {459}	0.999495	0.000044	0.003006	0.010020	OK.	1.000000	0.000000	1	0
Sn 189.989 {477}	0.999929	0.000006	0.001139	0.003797	OK.	1.000000	0.000000	1	0
Sr 407.771 { 83}	0.999996	0.000074	0.000554	0.001846	OK.	1.000000	0.000000	1	0
Ti 337.280 {100}	0.999177	0.000035	0.008833	0.029443	OK.	1.000000	0.000000	1	0
Tl 190.856 {477}	0.999963	0.000001	0.003985	0.013284	OK.	1.000000	0.000000	1	0
V 292.402 {115}	0.999980	0.000003	0.001255	0.004184	OK.	1.000000	0.000000	1	0
Y 224.306 {450}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Y 360.073 { 94}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Y 377.433 { 89}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Zn 206.200 {463}	0.999976	0.000021	0.000258	0.000861	OK.	1.000000	0.000000	1	0
Zr 339.198 { 99}	0.347379	0.000119	0.361260	1.204198	OK.	1.000000	0.000000	1	0

Sample Name: S0 Acquired: 6/23/2016 16:29:39 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00029**	**.00129**	**-.00008**	**.00004**	**.03153**	**.00093**	**.00028**
Stddev	.00015	.00002	.00014	.00006	.00240	.00003	.00198
%RSD	52.623	1.3157	185.14	142.12	7.6014	3.6422	698.55
#1	-.00011	.00130	.00005	.00006	.03428	.00091	-.00196
#2	-.00038	.00127	-.00005	.00008	.03039	.00092	.00105
#3	-.00038	.00130	-.00023	-.00002	.02991	.00097	.00176

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00014**	**-.00025**	**.00037**	**-.00034**	**-.00049**	**-.01061**	**-.01182**
Stddev	.00021	.00007	.00007	.00028	.00069	.00571	.01080
%RSD	147.13	28.247	17.769	80.896	139.88	53.794	91.363
#1	-.00024	-.00033	.00043	-.00059	.00004	-.01711	-.01999
#2	.00010	-.00020	.00030	-.00004	-.00127	-.00832	.00042
#3	-.00029	-.00021	.00036	-.00039	-.00024	-.00641	-.01590

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00070**	**-.00028**	**.00005**	**-.06454**	**-.00094**	**-.00026**	**-.00068**
Stddev	.00069	.00044	.00010	.00127	.00021	.00003	.00023
%RSD	98.775	157.73	191.66	1.9745	22.162	11.455	33.344
#1	-.00130	.00016	-.00002	-.06443	-.00114	-.00027	-.00091
#2	-.00086	-.00028	.00017	-.06587	-.00072	-.00029	-.00046
#3	.00006	-.00072	.00001	-.06333	-.00096	-.00023	-.00067

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00001**	**-.00025**	**.00251**	**.00006**	**.00245**	**-.00303**	**-.00041**
Stddev	.00016	.00016	.00016	.00001	.00241	.00117	.00012
%RSD	1654.9	64.234	6.3986	13.348	98.630	38.681	28.019
#1	-.00002	-.00035	.00248	.00005	.00157	-.00260	-.00029
#2	-.00014	-.00034	.00237	.00007	.00059	-.00213	-.00042
#3	.00019	-.00007	.00269	.00006	.00518	-.00435	-.00053

Sample Name: S0 Acquired: 6/23/2016 16:29:39 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S
Avg	.00000	.00017	-.01999
Stddev	.00010	.00005	.00203
%RSD	13188.	28.598	10.137
#1	.00004	.00021	-.01773
#2	-.00011	.00012	-.02164
#3	.00008	.00017	-.02061

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6699.4	41370.	2127.3
Stddev	40.4	297.	43.0
%RSD	.60365	.71762	2.0194
#1	6718.4	41699.	2078.6
#2	6726.8	41290.	2159.8
#3	6653.0	41122.	2143.5

Sample Name: S1 Acquired: 6/23/2016 16:33:38 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	Ba4554	Be3131	Ca4226	Cd2288	Co2286
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00002**	**.00207**	**.04860**	**.00138**	**.00433**	**.00022**	**.00042**
Stddev	.00014	.00009	.00373	.00005	.00188	.00012	.00011
%RSD	835.38	4.5773	7.6793	3.9714	43.525	57.028	26.394
#1	-.00012	.00203	.04518	.00133	.00557	.00008	.00054
#2	-.00007	.00218	.05258	.00138	.00216	.00025	.00040
#3	.00015	.00200	.04803	.00144	.00525	.00032	.00032

Elem	Cr2677	Cu2247	Fe2611	K_7664	Mn2576	Mo2020	Na5895
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00058**	**.00026**	**.00071**	**.01834**	**.00157**	**.00165**	**.02022**
Stddev	.00008	.00010	.00029	.00177	.00050	.00017	.00883
%RSD	13.891	39.383	40.273	9.6501	31.586	10.184	43.651
#1	.00049	.00020	.00042	.02032	.00204	.00159	.01736
#2	.00061	.00038	.00072	.01689	.00105	.00184	.01318
#3	.00064	.00021	.00099	.01782	.00163	.00152	.03013

Elem	Ni2316	P_2149	Pb2203	Sb2068	Si2124	Sn1899	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00007**	**.00087**	**-.00042**	**.00040**	**.00339**	**.00056**	**.03432**
Stddev	.00012	.00001	.00016	.00029	.00012	.00009	.00173
%RSD	168.70	1.1803	37.487	70.628	3.5276	16.358	5.0547
#1	-.00003	.00088	-.00031	.00019	.00333	.00066	.03349
#2	-.00020	.00086	-.00060	.00073	.00330	.00055	.03631
#3	.00002	.00087	-.00035	.00029	.00352	.00048	.03315

Elem	Ti3372	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00279**	**.00063**	**.00386**	**-.01703**
Stddev	.00101	.00006	.00004	.00240
%RSD	36.246	8.7705	.91033	14.105
#1	-.00167	.00068	.00390	-.01827
#2	-.00364	.00064	.00384	-.01856
#3	-.00305	.00057	.00384	-.01426

Sample Name: S1 Acquired: 6/23/2016 16:33:38 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6745.7**	**41805.**	**2129.8**
Stddev	22.3	250.	32.4
%RSD	.33088	.59789	1.5201
#1	6735.5	41617.	2128.0
#2	6730.4	42089.	2098.4
#3	6771.4	41710.	2163.1

Sample Name: S2 Acquired: 6/23/2016 16:37:37 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00004**	**.00279**	**.00016**	**.00026**	**.06973**	**.00179**	**.00735**
Stddev	.00008	.00007	.00007	.00002	.00127	.00005	.00035
%RSD	185.58	2.6503	44.732	8.8024	1.8178	2.7546	4.6908
#1	-.00013	.00287	.00008	.00023	.06878	.00184	.00697
#2	-.00000	.00273	.00022	.00027	.07117	.00174	.00747
#3	.00001	.00276	.00018	.00027	.06924	.00179	.00763

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00048**	**.00123**	**.00088**	**.00073**	**.00097**	**.04869**	**.00987**
Stddev	.00017	.00004	.00007	.00007	.00080	.01177	.00041
%RSD	36.245	2.8717	7.5775	10.086	82.497	24.183	4.1595
#1	.00028	.00126	.00081	.00081	.00163	.05564	.00947
#2	.00060	.00124	.00089	.00066	.00008	.03509	.01029
#3	.00056	.00119	.00095	.00072	.00119	.05534	.00985

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00079**	**.00313**	**.00327**	**.09617**	**.00018**	**.00208**	**-.00005**
Stddev	.00042	.00021	.00008	.00423	.00014	.00007	.00012
%RSD	53.033	6.7962	2.3656	4.4010	80.620	3.1401	235.98
#1	.00120	.00337	.00326	.09640	.00034	.00201	.00009
#2	.00080	.00296	.00319	.09182	.00013	.00215	-.00010
#3	.00036	.00306	.00335	.10028	.00006	.00208	-.00013

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00074**	**-.00014**	**.00523**	**.00123**	**.06780**	**-.00103**	**-.00017**
Stddev	.00018	.00007	.00014	.00004	.00229	.00100	.00009
%RSD	24.515	47.215	2.6523	3.3445	3.3781	97.203	54.232
#1	.00091	-.00021	.00538	.00120	.06958	-.00212	-.00021
#2	.00075	-.00013	.00511	.00121	.06861	-.00081	-.00007
#3	.00055	-.00008	.00519	.00127	.06522	-.00015	-.00025

Sample Name: S2 Acquired: 6/23/2016 16:37:37 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S
Avg	**.00124**	**.00773**	**-.01963**
Stddev	.00006	.00011	.00107
%RSD	4.9778	1.3726	5.4480
#1	.00125	.00769	-.01940
#2	.00117	.00765	-.01869
#3	.00129	.00785	-.02079
Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6707.0**	**42016.**	**2132.5**
Stddev	31.7	150.	7.6
%RSD	.47214	.35787	.35792
#1	6673.0	41936.	2136.2
#2	6712.3	41922.	2137.6
#3	6735.6	42189.	2123.7

Sample Name: S3 Acquired: 6/23/2016 16:41:36 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.02256**	**.09496**	**.01071**	**.01256**	**2.3852**	**.05519**	**.49691**
Stddev	.00004	.00014	.00002	.00013	.0192	.00031	.00535
%RSD	.17307	.15212	.16075	1.0212	.80533	.56599	1.0768
#1	.02258	.09479	.01070	.01253	2.3633	.05545	.49075
#2	.02251	.09505	.01073	.01245	2.3992	.05484	.50038
#3	.02258	.09504	.01070	.01270	2.3931	.05527	.49960

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.03562**	**.08233**	**.03402**	**.07207**	**.09984**	**3.4482**	**1.6795**
Stddev	.00009	.00032	.00019	.00009	.00122	.0308	.0141
%RSD	.25272	.38620	.54823	.12427	1.2238	.89404	.83921
#1	.03572	.08197	.03404	.07201	.09875	3.4126	1.6633
#2	.03558	.08257	.03383	.07204	.09961	3.4653	1.6862
#3	.03556	.08245	.03420	.07218	.10116	3.4666	1.6890

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.09604**	**.15811**	**.20462**	**9.6297**	**.07778**	**.14758**	**.03484**
Stddev	.00184	.00180	.00109	.0816	.00017	.00014	.00035
%RSD	1.9192	1.1373	.53383	.84731	.21367	.09481	1.0023
#1	.09394	.15608	.20545	9.5375	.07784	.14750	.03445
#2	.09678	.15875	.20503	9.6927	.07791	.14775	.03512
#3	.09739	.15949	.20338	9.6588	.07759	.14751	.03495

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.04088**	**.00687**	**.22552**	**.07884**	**3.9495**	**.13075**	**.01457**
Stddev	.00018	.00008	.00024	.00017	.0426	.00233	.00002
%RSD	.43951	1.1502	.10703	.21290	1.0795	1.7820	.16844
#1	.04086	.00680	.22528	.07865	3.9016	.12806	.01460
#2	.04106	.00686	.22576	.07898	3.9835	.13221	.01455
#3	.04070	.00696	.22552	.07889	3.9633	.13197	.01456

Sample Name: S3 Acquired: 6/23/2016 16:41:36 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S
Avg	**.07955**	**.47276**	**-.01302**
Stddev	.00051	.00054	.00096
%RSD	.63790	.11390	7.4014
#1	.07969	.47336	-.01283
#2	.07898	.47259	-.01406
#3	.07996	.47232	-.01217
Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6638.5**	**41219.**	**2121.0**
Stddev	46.3	234.	20.8
%RSD	.69679	.56802	.98295
#1	6589.4	41153.	2142.8
#2	6644.8	41479.	2101.2
#3	6681.3	41025.	2119.2

Sample Name: S4 Acquired: 6/23/2016 16:45:14 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.04536**	**.18763**	**.02145**	**.02504**	**4.7264**	**.10809**	**.99272**
Stddev	.00018	.00035	.00006	.00014	.0350	.00038	.00548
%RSD	.40723	.18482	.29786	.54138	.74034	.35140	.55236
#1	.04517	.18737	.02151	.02493	4.6879	.10794	.98645
#2	.04537	.18748	.02147	.02519	4.7563	.10781	.99661
#3	.04554	.18802	.02139	.02499	4.7351	.10852	.99512

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.07067**	**.16242**	**.06673**	**.14133**	**.20036**	**6.9280**	**3.3265**
Stddev	.00015	.00038	.00017	.00047	.00154	.0408	.0160
%RSD	.20929	.23179	.25203	.33195	.76659	.58859	.48243
#1	.07077	.16232	.06654	.14092	.19934	6.8881	3.3089
#2	.07075	.16283	.06687	.14184	.20213	6.9696	3.3403
#3	.07050	.16210	.06677	.14122	.19962	6.9263	3.3303

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.19280**	**.31594**	**.40461**	**19.378**	**.15355**	**.29587**	**.06972**
Stddev	.00166	.00297	.00142	.119	.00024	.00033	.00055
%RSD	.85870	.94051	.34987	.61191	.15345	.11299	.78837
#1	.19091	.31269	.40561	19.253	.15352	.29579	.06923
#2	.19352	.31659	.40524	19.489	.15380	.29623	.07032
#3	.19397	.31853	.40299	19.392	.15333	.29558	.06962

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.08150**	**.01388**	**.44160**	**.15662**	**7.8766**	**.26694**	**.02902**
Stddev	.00043	.00013	.00090	.00039	.0522	.00276	.00015
%RSD	.53112	.92390	.20356	.24938	.66265	1.0329	.50476
#1	.08128	.01397	.44144	.15669	7.8213	.26376	.02890
#2	.08200	.01394	.44257	.15697	7.9249	.26844	.02918
#3	.08123	.01374	.44079	.15620	7.8835	.26861	.02897

Sample Name: S4 Acquired: 6/23/2016 16:45:14 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S
Avg	**.15703**	**.93159**	**-.00927**
Stddev	.00036	.00158	.00064
%RSD	.22618	.16981	6.9563
#1	.15712	.93104	-.00858
#2	.15664	.93337	-.00938
#3	.15734	.93035	-.00985

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6531.9**	**40084.**	**2097.6**
Stddev	16.2	231.	6.3
%RSD	.24773	.57740	.30149
#1	6549.4	40239.	2103.9
#2	6517.6	40195.	2091.2
#3	6528.6	39818.	2097.7

Sample Name: ICV Acquired: 6/23/2016 16:48:51 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39953	10.024	.40138	.50599	1.0070	.05053	10.036
Stddev	.00339	.040	.00410	.00222	.0059	.00022	.101
%RSD	.84874	.40015	1.0204	.43801	.58442	.43218	1.0067
#1	.40015	10.065	.39887	.50853	1.0016	.05072	9.9340
#2	.39587	9.9844	.39916	.50451	1.0063	.05029	10.037
#3	.40256	10.022	.40610	.50491	1.0133	.05057	10.136
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05000	.20142	.50405	.50578	4.0103	50.167	1.0145
Stddev	.00018	.00074	.00219	.00236	.0323	.251	.0072
%RSD	.36119	.36866	.43440	.46709	.80429	.50069	.71348
#1	.04994	.20142	.50464	.50308	4.0274	49.888	1.0064
#2	.05021	.20217	.50162	.50746	3.9731	50.374	1.0203
#3	.04986	.20068	.50588	.50680	4.0305	50.241	1.0167
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	10.075	.50106	1.0146	50.102	.50246	9.9799	.50553
Stddev	.156	.00520	.0033	.232	.00089	.0190	.00257
%RSD	1.5448	1.0383	.32210	.46381	.17782	.18993	.50874
#1	9.9213	.49528	1.0136	49.851	.50217	9.9887	.50315
#2	10.071	.50252	1.0182	50.148	.50346	9.9928	.50825
#3	10.233	.50537	1.0119	50.309	.50174	9.9581	.50518
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: ICV Acquired: 6/23/2016 16:48:51 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2034**	**.39904**	**5.0662**	**1.0017**	**1.0068**	**.99368**	**.50263**
Stddev	.0053	.01223	.0118	.0027	.0059	.01682	.00169
%RSD	.44447	3.0652	.23394	.26994	.59001	1.6930	.33671
#1	1.2003	.41278	5.0644	1.0010	1.0005	.98203	.50067
#2	1.2095	.38936	5.0789	1.0047	1.0076	.98605	.50350
#3	1.2003	.39496	5.0554	.99945	1.0123	1.0130	.50370
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0106**	**1.0101**	F **.08377**
Stddev	.0013	.0017	.57243
%RSD	.13014	.17012	683.32
#1	1.0108	1.0096	.41336
#2	1.0092	1.0121	.41517
#3	1.0118	1.0088	-.57721
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-5.0000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6582.9**	**40798.**	**2104.4**
Stddev	57.1	288.	9.3
%RSD	.86719	.70695	.43973
#1	6642.0	40517.	2109.9
#2	6528.1	41093.	2109.6
#3	6578.7	40785.	2093.7

Sample Name: ICB Acquired: 6/23/2016 16:52:29 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00205	.00812	.00204	.00053	.00042	.00004	.08043	.00000
Stddev	.00123	.00380	.00409	.00171	.00048	.00007	.04843	.00019
%RSD	60.002	46.772	200.40	325.30	114.02	191.36	60.215	12542.
#1	.00117	.01182	.00568	-.00071	.00071	-.00005	.13570	.00020
#2	.00153	.00832	-.00239	.00248	.00068	.00009	.04544	-.00018
#3	.00346	.00423	.00283	-.00019	-.00013	.00007	.06015	-.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00045	-.00032	.00256	.01917	.26687	-.00090	-.01824	.00334
Stddev	.00068	.00118	.00062	.01550	.00597	.00344	.19098	.00120
%RSD	152.08	365.07	24.256	80.873	2.2373	382.34	1047.0	35.896
#1	.00109	.00091	.00301	.01950	.26758	.00158	-.16456	.00207
#2	-.00027	-.00043	.00283	.00350	.26057	.00055	-.08796	.00445
#3	.00052	-.00145	.00185	.03450	.27245	-.00483	.19780	.00349
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00543	.02836	-.00115	.00608	-.00046	.00498	-.00449	-.00458
Stddev	.00101	.00804	.00072	.01165	.00093	.00238	.00576	.00313
%RSD	18.519	28.338	62.364	191.75	202.82	47.775	128.25	68.440
#1	.00443	.02219	-.00190	-.00553	.00045	.00510	-.01053	-.00111
#2	.00542	.03745	-.00110	.00598	-.00041	.00730	-.00387	-.00541
#3	.00644	.02544	-.00046	.01778	-.00141	.00254	.00094	-.00721
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: ICB Acquired: 6/23/2016 16:52:29 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00002	.00059	-.00321	-.00285	-.00021	.00049	-.01967
Stddev	.00049	.00058	.00271	.00281	.00126	.00023	.30758
%RSD	1987.7	98.883	84.402	98.601	601.40	48.028	1563.4
#1	.00048	.00125	-.00307	-.00514	.00122	.00050	-.07765
#2	-.00050	.00018	-.00600	.00028	-.00069	.00072	.31277
#3	.00010	.00032	-.00058	-.00369	-.00116	.00025	-.29414
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6667.4	41423.	2115.8
Stddev	41.6	269.	6.3
%RSD	.62409	.64899	.29744
#1	6626.3	41592.	2109.7
#2	6666.3	41563.	2115.3
#3	6709.5	41113.	2122.3

Sample Name: LLICV Acquired: 6/23/2016 16:56:17 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00827	.16210	.00867	.07876	.00792	.00155	.45064	.00085
Stddev	.00069	.00016	.00152	.00325	.00022	.00012	.03526	.00030
%RSD	8.3443	.10048	17.555	4.1248	2.8338	7.9064	7.8234	35.681
#1	.00876	.16217	.00804	.07520	.00817	.00167	.42049	.00097
#2	.00858	.16223	.00757	.07953	.00782	.00155	.48940	.00050
#3	.00748	.16192	.01041	.08156	.00775	.00143	.44202	.00107
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00411	.00286	.00483	.07579	.81519	.07605	.43031	.01159
Stddev	.00051	.00122	.00183	.02010	.08172	.00054	.07836	.00240
%RSD	12.383	42.665	37.852	26.521	10.025	.70809	18.211	20.716
#1	.00446	.00209	.00311	.08314	.72240	.07546	.51973	.01210
#2	.00353	.00427	.00463	.05305	.87645	.07652	.37359	.01370
#3	.00433	.00223	.00676	.09119	.84672	.07616	.39761	.00898
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00886	.44734	.01615	.77513	.00676	.08645	.01136	.83408
Stddev	.00038	.03636	.00114	.01025	.00094	.00237	.00475	.00045
%RSD	4.3027	8.1283	7.0552	1.3217	13.970	2.7462	41.818	.05401
#1	.00863	.41320	.01707	.76914	.00779	.08743	.00922	.83393
#2	.00866	.44325	.01488	.78696	.00652	.08374	.01680	.83373
#3	.00930	.48558	.01652	.76928	.00595	.08818	.00805	.83459
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: LLICV Acquired: 6/23/2016 16:56:17 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.39937**	**.04063**	**.02339**	**.14972**	**.00726**	**.01640**	**20.158**
Stddev	.00149	.00022	.00849	.00368	.00013	.00008	.057
%RSD	.37294	.54009	36.301	2.4588	1.7938	.47839	.28459
#1	.40039	.04052	.01395	.15061	.00733	.01635	20.157
#2	.40005	.04089	.02583	.15288	.00711	.01649	20.101
#3	.39766	.04050	.03040	.14568	.00735	.01636	20.216
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6699.9**	**41507.**	**2122.8**
Stddev	10.7	321.	33.0
%RSD	.15908	.77425	1.5534
#1	6695.4	41349.	2085.0
#2	6692.2	41295.	2145.6
#3	6712.1	41877.	2137.8

Sample Name: LLICV Acquired: 6/23/2016 17:00:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01949	.39961	.01957	.19216	.01939	.00378	1.0069
Stddev	.00131	.00492	.00179	.00362	.00127	.00002	.0595
%RSD	6.7128	1.2306	9.1680	1.8843	6.5460	.43946	5.9133
#1	.02095	.40406	.02038	.19600	.02017	.00376	.96045
#2	.01841	.40043	.01752	.18881	.01792	.00378	.98624
#3	.01913	.39433	.02082	.19166	.02007	.00379	1.0740
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00195	.01011	.00920	.01111	.21788	2.0711	.19482
Stddev	.00022	.00045	.00124	.00158	.02611	.1150	.00581
%RSD	11.156	4.4276	13.477	14.191	11.985	5.5533	2.9817
#1	.00196	.01037	.01057	.01272	.21124	2.1436	.18839
#2	.00173	.01037	.00885	.00957	.19573	2.1311	.19636
#3	.00216	.00960	.00817	.01104	.24667	1.9385	.19970
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.0302	.01992	.01953	.99004	.04110	1.9033	.01787
Stddev	.1225	.00176	.00044	.02149	.00123	.0089	.00330
%RSD	11.893	8.8312	2.2495	2.1708	2.9826	.46494	18.472
#1	.89819	.01975	.02003	.96810	.03968	1.8971	.02156
#2	1.0522	.01825	.01927	1.0110	.04177	1.9135	.01518
#3	1.1403	.02176	.01928	.99098	.04184	1.8995	.01689
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: LLICV Acquired: 6/23/2016 17:00:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20779	.03211	2.0518	.97692	.09846	.05795	.36369
Stddev	.00270	.00309	.0013	.00240	.00149	.00439	.00103
%RSD	1.2971	9.6240	.06110	.24594	1.5132	7.5738	.28252
#1	.21089	.03387	2.0505	.97555	.09739	.05293	.36373
#2	.20647	.02854	2.0530	.97969	.09783	.06107	.36264
#3	.20602	.03392	2.0519	.97551	.10016	.05984	.36469
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.01983	.04054	F 52.666
Stddev	.00096	.00030	.271
%RSD	4.8538	.74967	.51545
#1	.01995	.04039	52.957
#2	.01882	.04089	52.622
#3	.02073	.04034	52.420
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6722.9	41992.	2132.9
Stddev	4.6	356.	26.2
%RSD	.06888	.84767	1.2278
#1	6727.0	42116.	2114.2
#2	6723.9	42269.	2162.8
#3	6717.8	41590.	2121.7

Sample Name: ICSA Acquired: 6/23/2016 17:04:08 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00149	266.38	-.00372	.01062	.00153	.00000	253.32
Stddev	.00133	.59	.00223	.00385	.00251	.00008	1.14
%RSD	89.471	.22308	59.913	36.253	163.63	4300.4	.45010
#1	-.00003	267.06	-.00307	.00628	.00113	-.00004	252.01
#2	.00202	266.01	-.00189	.01196	-.00075	-.00004	253.84
#3	.00246	266.05	-.00620	.01362	.00422	.00009	254.10
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00047	-.00114	-.00174	.00303	99.707	.16334	.00688
Stddev	.00038	.00002	.00038	.00035	.605	.10414	.00241
%RSD	79.353	1.7935	21.793	11.497	.60634	63.757	34.985
#1	.00088	-.00115	-.00217	.00298	99.093	.28269	.00955
#2	.00015	-.00116	-.00144	.00340	99.728	.09096	.00486
#3	.00039	-.00112	-.00162	.00271	100.30	.11636	.00625
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	252.57	.00029	.00006	.04525	-.00141	.04327	-.00195
Stddev	.76	.00143	.00058	.00703	.00150	.00827	.00429
%RSD	.29957	489.02	1021.2	15.529	106.39	19.108	220.24
#1	251.84	.00193	-.00061	.05054	-.00123	.05059	-.00454
#2	252.51	-.00031	.00047	.04793	-.00001	.04493	.00300
#3	253.35	-.00074	.00031	.03728	-.00299	.03430	-.00431
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: ICSA Acquired: 6/23/2016 17:04:08 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00596**	**-.00219**	**-.00062**	**.00088**	**.00057**	**-.00643**	**.00061**
Stddev	.00409	.01112	.00241	.00133	.00012	.00454	.00641
%RSD	68.491	508.56	387.55	152.27	20.327	70.638	1044.9
#1	-.01030	-.01132	.00215	.00215	.00049	-.00321	.00778
#2	-.00542	-.00543	-.00218	-.00050	.00052	-.00446	-.00133
#3	-.00218	.01019	-.00184	.00098	.00071	-.01163	-.00460
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00692**	**-.00188**	F **-.35802**
Stddev	.00157	.00029	.22129
%RSD	22.711	15.411	61.810
#1	-.00512	-.00157	-.10687
#2	-.00763	-.00196	-.44284
#3	-.00802	-.00213	-.52437
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.02000
Low Limit			-.02000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6250.1**	**37849.**	**2031.8**
Stddev	54.8	56.	7.8
%RSD	.87682	.14906	.38488
#1	6190.3	37786.	2027.2
#2	6297.9	37868.	2027.4
#3	6262.0	37894.	2040.9

Sample Name: ICSAB Acquired: 6/23/2016 17:08:01 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.52882	265.17	.25293	.00992	.25484	.24777	252.21
Stddev	.00098	.76	.00285	.00453	.00061	.00074	1.18
%RSD	.18558	.28502	1.1258	45.657	.23964	.29732	.46623
#1	.52981	264.89	.25009	.01062	.25418	.24751	252.99
#2	.52881	266.03	.25579	.00508	.25539	.24861	250.86
#3	.52785	264.59	.25291	.01406	.25495	.24721	252.79
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50186	.24161	.25459	.25471	97.424	5.2402	.00666
Stddev	.00151	.00122	.00264	.00209	.509	.0634	.00798
%RSD	.30055	.50701	1.0371	.81915	.52250	1.2106	119.82
#1	.50037	.24023	.25660	.25353	97.770	5.2868	.00639
#2	.50338	.24256	.25160	.25711	96.840	5.1680	.01477
#3	.50184	.24204	.25558	.25347	97.663	5.2658	-.00118
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	247.66	.24890	-.00112	5.3746	.48553	.04636	.50217
Stddev	1.37	.00098	.00019	.0121	.00314	.00695	.00280
%RSD	.55427	.39569	17.374	.22614	.64678	14.982	.55722
#1	248.16	.24818	-.00127	5.3707	.48197	.04411	.50327
#2	246.11	.25002	-.00118	5.3648	.48669	.04082	.50425
#3	248.71	.24850	-.00090	5.3882	.48792	.05415	.49899
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: ICSAB Acquired: 6/23/2016 17:08:01 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50735	.24194	-.00758	-.00008	.00043	-.00758	.46915
Stddev	.00346	.00285	.00236	.00064	.00025	.00399	.00306
%RSD	.68179	1.1790	31.189	795.11	59.086	52.651	.65264
#1	.51111	.24253	-.00918	-.00025	.00026	-.00923	.46976
#2	.50431	.23884	-.00486	.00062	.00031	-.00303	.46582
#3	.50662	.24445	-.00869	-.00061	.00072	-.01049	.47185
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.25055	.48484	F -.54156
Stddev	.00015	.00117	.36388
%RSD	.05947	.24090	67.191
#1	.25062	.48352	-.86893
#2	.25038	.48574	-.14978
#3	.25065	.48526	-.60595
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.02500
Low Limit			-.02500

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6244.9	38520.	2058.8
Stddev	46.1	338.	32.6
%RSD	.73809	.87856	1.5823
#1	6288.3	38547.	2028.4
#2	6196.5	38169.	2093.1
#3	6249.8	38844.	2054.9

Sample Name: CCV Acquired: 6/23/2016 17:11:45 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39852	10.052	.40263	.49843	1.0023	.05021	10.092
Stddev	.00333	.028	.00314	.00166	.0006	.00016	.057
%RSD	.83575	.27814	.78076	.33361	.06012	.31593	.56540
#1	.39689	10.041	.39901	.49807	1.0017	.05007	10.131
#2	.40235	10.030	.40421	.49698	1.0029	.05018	10.027
#3	.39632	10.083	.40467	.50024	1.0023	.05038	10.118
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04999	.20043	.50163	.50411	4.0601	50.005	1.0189
Stddev	.00015	.00042	.00302	.00115	.0027	.046	.0052
%RSD	.29315	.20823	.60186	.22714	.06779	.09107	.50520
#1	.05012	.20084	.49900	.50511	4.0627	50.032	1.0215
#2	.04983	.20000	.50096	.50286	4.0604	49.953	1.0130
#3	.05001	.20044	.50492	.50436	4.0572	50.031	1.0222
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	10.212	.49892	1.0009	49.902	.50256	9.9305	.50524
Stddev	.091	.00599	.0036	.189	.00114	.0123	.00430
%RSD	.88794	1.1999	.36370	.37915	.22664	.12410	.85170
#1	10.107	.49738	1.0047	49.717	.50312	9.9163	.50068
#2	10.266	.49386	1.0004	50.095	.50125	9.9365	.50580
#3	10.263	.50553	.99747	49.894	.50331	9.9386	.50923
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 17:11:45 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1976**	**.39739**	**5.0461**	**.99886**	**.99946**	**1.0004**	**.50205**
Stddev	.0044	.00576	.0095	.00286	.00488	.0097	.00380
%RSD	.36569	1.4500	.18873	.28629	.48793	.96796	.75754
#1	1.2025	.39818	5.0436	1.0005	.99383	.99232	.49767
#2	1.1964	.39127	5.0380	1.0005	1.0021	.99772	.50399
#3	1.1940	.40272	5.0566	.99556	1.0025	1.0111	.50449
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0051**	**1.0042**	F **.84890**
Stddev	.0013	.0014	.43937
%RSD	.13211	.13734	51.758
#1	1.0042	1.0051	.99034
#2	1.0045	1.0026	.35623
#3	1.0066	1.0048	1.2001
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6596.3**	**41288.**	**2118.4**
Stddev	44.5	204.	20.3
%RSD	.67419	.49375	.95799
#1	6545.1	41449.	2095.9
#2	6618.2	41357.	2123.7
#3	6625.5	41059.	2135.5

Sample Name: CCB Acquired: 6/23/2016 17:15:23 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00313**	**.01723**	**.00317**	**.00223**	**-.00086**	**.00005**	**.05209**
Stddev	.00122	.00350	.00117	.00121	.00076	.00008	.06012
%RSD	38.963	20.296	36.752	54.255	88.561	155.39	115.43
#1	.00430	.02036	.00417	.00084	-.00098	.00005	.05688
#2	.00322	.01787	.00345	.00279	-.00004	.00013	.10967
#3	.00186	.01346	.00189	.00306	-.00155	-.00003	-.01029
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**.00041**	**.00080**	**.00088**	**.00933**	**.12101**	**-.00019**
Stddev	.00009	.00032	.00113	.00031	.02938	.20550	.00178
%RSD	112.14	77.747	141.29	34.787	314.94	169.82	929.52
#1	-.00002	.00017	.00160	.00098	.03104	-.04316	-.00153
#2	.00015	.00030	.00129	.00054	-.02411	.35147	-.00087
#3	.00013	.00078	-.00049	.00114	.02106	.05472	.00183
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.08095**	**.00318**	**.00354**	**.07516**	**-.00052**	**.00211**	**-.00169**
Stddev	.03617	.00076	.00074	.03642	.00171	.00399	.00228
%RSD	44.681	24.034	20.829	48.460	332.21	188.88	134.38
#1	.11808	.00257	.00275	.11180	-.00209	.00394	.00084
#2	.04583	.00292	.00421	.07473	-.00077	-.00247	-.00235
#3	.07893	.00403	.00365	.03896	.00131	.00486	-.00357
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 17:15:23 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00292**	**.00064**	**-.00112**	**.00037**	**.00055**	**.00009**	**-.00222**
Stddev	.00495	.00211	.00181	.00085	.00039	.00846	.00125
%RSD	169.66	329.79	161.50	228.24	71.494	9095.7	56.323
#1	.00726	.00025	-.00053	.00012	.00049	.00531	-.00357
#2	.00396	.00292	-.00315	.00131	.00097	-.00967	-.00109
#3	-.00247	-.00125	.00032	-.00032	.00019	.00464	-.00201
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00045**	**.00017**	F **-.31880**
Stddev	.00139	.00012	.06794
%RSD	311.50	72.605	21.311
#1	-.00109	.00003	-.30839
#2	.00115	.00026	-.39134
#3	-.00140	.00021	-.25666
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6671.9**	**41329.**	**2107.9**
Stddev	60.0	164.	15.1
%RSD	.89934	.39564	.71641
#1	6603.0	41427.	2098.9
#2	6712.5	41420.	2099.5
#3	6700.3	41140.	2125.4

Sample Name: PBW LJ Acquired: 6/23/2016 17:19:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00080**	**.01261**	**.00190**	**.00491**	**-.00092**	**-.00002**	**.05100**	**.00023**
Stddev	.00084	.00763	.00125	.00212	.00119	.00006	.03078	.00026
%RSD	103.96	60.561	65.725	43.302	129.08	259.98	60.367	110.85
#1	.00134	.02138	.00270	.00692	-.00108	-.00003	.01560	.00021
#2	-.00016	.00898	.00254	.00512	.00034	.00004	.07152	-.00001
#3	.00123	.00746	.00046	.00268	-.00203	-.00008	.06587	.00050
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00037**	**-.00004**	**.00109**	**.01291**	**.08713**	**-.00094**	**.00517**	**.00144**
Stddev	.00024	.00123	.00204	.01662	.13771	.00409	.05412	.00202
%RSD	65.746	3083.5	187.27	128.71	158.06	434.61	1047.5	139.96
#1	.00019	-.00005	-.00097	-.00552	-.06256	.00328	.06512	-.00079
#2	.00064	-.00127	.00113	.02675	.20845	-.00488	-.00956	.00312
#3	.00026	.00119	.00311	.01750	.11550	-.00122	-.04006	.00199
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00095**	**.04152**	**-.00019**	**-.00390**	**-.00371**	**.00030**	**.00094**	**-.01516**
Stddev	.00059	.04832	.00205	.00634	.00397	.00313	.00086	.00318
%RSD	62.354	116.37	1063.0	162.76	106.95	1035.6	91.391	20.967
#1	.00033	.01830	-.00241	-.00387	-.00717	.00089	.00151	-.01165
#2	.00102	.00919	.00019	.00243	.00062	.00310	.00135	-.01599
#3	.00151	.09706	.00164	-.01025	-.00458	-.00308	-.00005	-.01784
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: PBW LJ Acquired: 6/23/2016 17:19:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00033**	**.00039**	**-.00004**	**-.00270**	**-.00095**	**.00140**	**.41056**
Stddev	.00027	.00036	.00305	.00089	.00074	.00017	.24431
%RSD	82.432	92.835	6840.9	33.104	77.766	12.399	59.505
#1	-.00063	.00081	-.00312	-.00373	-.00092	.00123	.69212
#2	-.00012	.00013	.00297	-.00211	-.00022	.00158	.28489
#3	-.00024	.00024	.00002	-.00226	-.00170	.00139	.25468
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6640.4**	**41625.**	**2128.5**
Stddev	17.6	211.	40.6
%RSD	.26566	.50582	1.9066
#1	6652.6	41596.	2081.7
#2	6648.4	41848.	2149.1
#3	6620.1	41430.	2154.5

Sample Name: LCSW LJ Acquired: 6/23/2016 17:23:21 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20319	5.0131	.20150	.99773	.51668	.02479	5.2442	.02487
Stddev	.00169	.0087	.00325	.00753	.00370	.00010	.0451	.00015
%RSD	.82977	.17310	1.6126	.75506	.71666	.41148	.86061	.60213
#1	.20357	5.0214	.19782	.99002	.52094	.02487	5.2832	.02481
#2	.20466	5.0041	.20272	1.0051	.51488	.02467	5.2546	.02476
#3	.20135	5.0138	.20397	.99808	.51422	.02483	5.1948	.02504
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10159	.25137	.25763	2.1220	25.452	.52645	5.1614	.25904
Stddev	.00032	.00052	.00140	.0264	.157	.00783	.0562	.00289
%RSD	.31802	.20500	.54330	1.2440	.61733	1.4875	1.0892	1.1168
#1	.10195	.25197	.25859	2.1525	25.633	.53504	5.1502	.26071
#2	.10149	.25109	.25602	2.1081	25.375	.52462	5.1116	.26070
#3	.10132	.25106	.25826	2.1055	25.349	.51970	5.2224	.25570
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.51626	25.724	.25483	4.8485	.25110	.61184	.19096	2.6019
Stddev	.00266	.216	.00018	.0108	.00443	.00393	.01056	.0098
%RSD	.51446	.83901	.07257	.22239	1.7625	.64191	5.5294	.37803
#1	.51860	25.969	.25482	4.8553	.24906	.61631	.19143	2.6112
#2	.51337	25.641	.25502	4.8360	.24806	.60893	.18017	2.5916
#3	.51679	25.562	.25465	4.8540	.25618	.61029	.20127	2.6028
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: LCSW LJ Acquired: 6/23/2016 17:23:21 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50721	.52436	.50961	.25007	.50364	.50657	.76573
Stddev	.00162	.00527	.01502	.00251	.00102	.00126	.62483
%RSD	.31955	1.0056	2.9465	1.0024	.20335	.24816	81.599
#1	.50743	.53024	.51039	.24815	.50410	.50766	1.3558
#2	.50550	.52279	.52423	.25290	.50247	.50519	.11116
#3	.50872	.52005	.49423	.24914	.50436	.50685	.83021
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6575.8	41413.	2125.8
Stddev	38.0	254.	15.3
%RSD	.57733	.61440	.72050
#1	6590.8	41169.	2116.8
#2	6532.6	41394.	2117.1
#3	6603.9	41677.	2143.4

Sample Name: L1606076404 Acquired: 6/23/2016 17:27:01 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00032**	**.03451**	**.00124**	**.00949**	**.00772**	**.00001**	**37.580**	**.00027**
Stddev	.00100	.00434	.00233	.00396	.00016	.00008	.209	.00010
%RSD	315.71	12.583	187.37	41.736	2.0214	559.86	.55655	38.752
#1	-.00025	.02981	-.00143	.01269	.00789	-.00008	37.790	.00037
#2	.00065	.03535	.00232	.00506	.00758	.00004	37.371	.00027
#3	-.00135	.03837	.00283	.01072	.00769	.00008	37.578	.00016
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00036**	**.01239**	**.00232**	**.45824**	**5.1190**	**.00322**	**14.570**	**.10315**
Stddev	.00021	.00017	.00134	.02324	.0359	.00604	.145	.00332
%RSD	57.485	1.3945	57.779	5.0714	.70061	187.38	.99354	3.2187
#1	.00054	.01232	.00335	.43843	5.0785	.00539	14.664	.10664
#2	.00040	.01226	.00280	.48382	5.1320	-.00360	14.403	.10004
#3	.00014	.01258	.00080	.45248	5.1466	.00788	14.643	.10276
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00112**	**35.286**	**.00502**	**.02785**	**.00193**	**.00400**	**.00017**	**4.2205**
Stddev	.00017	.168	.00085	.00292	.00187	.00379	.00613	.0101
%RSD	15.273	.47690	16.934	10.483	97.237	94.733	3643.4	.24042
#1	.00102	35.461	.00444	.02992	.00018	.00832	-.00425	4.2311
#2	.00102	35.126	.00463	.02451	.00170	.00241	-.00240	4.2196
#3	.00131	35.269	.00600	.02912	.00390	.00126	.00716	4.2109
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606076404 Acquired: 6/23/2016 17:27:01 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00009**	**.19838**	**-.00352**	**-.00316**	**-.00056**	**.00304**	**.05702**
Stddev	.00048	.00082	.00439	.00352	.00018	.00003	.32731
%RSD	532.68	.41544	124.63	111.42	32.096	1.0103	574.06
#1	-.00061	.19924	-.00146	-.00663	-.00039	.00305	.43214
#2	-.00001	.19830	-.00054	.00041	-.00055	.00306	-.17051
#3	.00035	.19760	-.00855	-.00325	-.00075	.00300	-.09057
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6659.9**	**41441.**	**2148.1**
Stddev	17.8	39.	34.6
%RSD	.26678	.09458	1.6087
#1	6640.6	41401.	2109.0
#2	6663.2	41442.	2174.8
#3	6675.7	41479.	2160.4

Sample Name: L1606076405S Acquired: 6/23/2016 17:30:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20092	5.0385	.20312	.99798	.52837	.02495	44.005	.02497
Stddev	.00159	.0071	.00090	.00399	.00863	.00018	.536	.00030
%RSD	.78894	.14165	.44224	.39989	1.6338	.73761	1.2192	1.2169
#1	.20053	5.0320	.20215	1.0021	.53714	.02516	44.468	.02467
#2	.20266	5.0374	.20392	.99408	.52809	.02480	44.130	.02527
#3	.19957	5.0461	.20329	.99779	.51988	.02489	43.417	.02498
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10016	.26480	.25157	2.5877	30.715	.52688	20.416	.36208
Stddev	.00067	.00097	.00149	.0124	.632	.01078	.244	.00320
%RSD	.66866	.36563	.59363	.47820	2.0592	2.0459	1.1951	.88399
#1	.09940	.26587	.24988	2.6017	31.379	.53733	20.633	.36533
#2	.10067	.26457	.25208	2.5782	30.645	.52751	20.464	.36196
#3	.10042	.26397	.25273	2.5832	30.120	.51580	20.152	.35894
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.52071	62.572	.25549	5.0021	.25301	.61345	.19142	7.0690
Stddev	.00097	.840	.00144	.0110	.00528	.00259	.00718	.0143
%RSD	.18659	1.3422	.56532	.21968	2.0876	.42179	3.7518	.20212
#1	.52096	63.421	.25626	4.9894	.25249	.61450	.18326	7.0650
#2	.52153	62.553	.25639	5.0088	.25853	.61534	.19422	7.0849
#3	.51964	61.741	.25383	5.0081	.24801	.61050	.19677	7.0571
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606076405S Acquired: 6/23/2016 17:30:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50671**	**.73344**	**.51400**	**.25052**	**.50766**	**.50196**	**.71683**
Stddev	.00049	.00764	.00562	.00109	.00170	.00107	.08607
%RSD	.09604	1.0422	1.0940	.43503	.33403	.21255	12.007
#1	.50635	.74103	.51985	.25089	.50822	.50082	.65957
#2	.50651	.73352	.50864	.24929	.50576	.50294	.81581
#3	.50726	.72575	.51349	.25138	.50901	.50210	.67512
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6617.0**	**40888.**	**2108.3**
Stddev	36.3	229.	43.4
%RSD	.54907	.56042	2.0607
#1	6597.9	40710.	2065.4
#2	6594.1	40808.	2107.2
#3	6658.9	41146.	2152.3

Sample Name: L1606076406SD Acquired: 6/23/2016 17:34:36 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20158	5.0081	.20119	.99289	.52520	.02487	44.269	.02478
Stddev	.00322	.0181	.00119	.00601	.00285	.00008	.382	.00037
%RSD	1.5962	.36106	.59360	.60498	.54186	.31907	.86303	1.4919
#1	.20051	5.0088	.20213	.99898	.52769	.02491	44.667	.02436
#2	.19904	5.0258	.20161	.99270	.52581	.02493	44.234	.02503
#3	.20520	4.9897	.19985	.98697	.52210	.02478	43.905	.02496
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09994	.26339	.24953	2.5877	30.662	.52181	20.366	.36158
Stddev	.00059	.00143	.00256	.0611	.222	.00474	.372	.00451
%RSD	.59271	.54249	1.0252	2.3607	.72504	.90810	1.8250	1.2464
#1	.09927	.26187	.24744	2.6536	30.906	.52560	20.793	.36652
#2	.10013	.26359	.24876	2.5330	30.471	.52334	20.120	.35770
#3	.10041	.26470	.25238	2.5764	30.609	.51650	20.184	.36051
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.51522	62.707	.25233	4.9723	.25332	.61176	.19301	7.0733
Stddev	.00118	.507	.00147	.0132	.00160	.00273	.00266	.0075
%RSD	.22851	.80791	.58315	.26561	.63178	.44609	1.3764	.10575
#1	.51652	63.283	.25117	4.9632	.25248	.61207	.19130	7.0646
#2	.51490	62.510	.25184	4.9664	.25231	.61431	.19166	7.0774
#3	.51423	62.328	.25399	4.9875	.25516	.60888	.19607	7.0778
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606076406SD Acquired: 6/23/2016 17:34:36 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50236	.73208	.50989	.24941	.50289	.49718	.41726
Stddev	.00239	.00589	.01162	.00247	.00231	.00067	.67397
%RSD	.47662	.80444	2.2794	.99175	.45970	.13482	161.52
#1	.49978	.73852	.52268	.24842	.50555	.49665	-.07543
#2	.50451	.73075	.49998	.25223	.50134	.49794	.14190
#3	.50280	.72697	.50700	.24759	.50179	.49697	1.1853
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6608.7	41166.	2120.3
Stddev	11.1	137.	19.6
%RSD	.16723	.33193	.92378
#1	6603.0	41323.	2097.8
#2	6621.5	41090.	2129.6
#3	6601.7	41083.	2133.5

Sample Name: L1606109801 Acquired: 6/23/2016 17:38:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00129**	**.01287**	**-.00287**	**.00614**	**-.00048**	**-.00003**	**.08304**	**.00009**
Stddev	.00159	.00811	.00192	.00213	.00074	.00007	.01337	.00013
%RSD	122.92	63.032	67.055	34.715	154.39	222.73	16.095	139.40
#1	.00212	.02216	-.00081	.00626	-.00008	-.00010	.09008	.00010
#2	.00231	.00925	-.00462	.00395	-.00134	.00003	.09142	-.00004
#3	-.00054	.00720	-.00317	.00821	-.00003	-.00001	.06763	.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00010**	**.00114**	**.00130**	**.02096**	**.10896**	**-.00006**	**-.03305**	**.00197**
Stddev	.00044	.00113	.00181	.01243	.08516	.00372	.07886	.00311
%RSD	453.08	99.110	139.52	59.304	78.156	6481.1	238.62	158.05
#1	-.00053	.00170	.00306	.02326	.19673	-.00420	-.07467	-.00106
#2	.00034	-.00016	-.00055	.03207	.02667	.00104	-.08238	.00515
#3	-.00009	.00187	.00138	.00754	.10349	.00299	.05790	.00182
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00066**	**.45527**	**.00049**	**-.00366**	**.00153**	**.00122**	**.00971**	**-.00748**
Stddev	.00030	.03584	.00013	.00895	.00348	.00363	.01502	.00042
%RSD	45.839	7.8733	25.556	244.66	227.48	297.89	154.66	5.5803
#1	.00053	.41497	.00054	-.01374	.00021	.00522	.02568	-.00782
#2	.00044	.46723	.00035	-.00059	-.00110	-.00186	-.00414	-.00701
#3	.00100	.48360	.00058	.00335	.00548	.00030	.00761	-.00761
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606109801 Acquired: 6/23/2016 17:38:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**.00480**	**-.00167**	**-.00415**	**-.00006**	**.00240**	**.05406**
Stddev	.00008	.00020	.00916	.00353	.00083	.00017	.09676
%RSD	178.62	4.1888	547.73	84.853	1386.4	7.2187	179.01
#1	-.00014	.00474	.00699	-.00318	.00001	.00255	.07032
#2	-.00003	.00462	-.01126	-.00807	-.00092	.00221	-.04981
#3	.00002	.00502	-.00075	-.00122	.00073	.00245	.14165
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6705.4**	**41704.**	**2110.9**
Stddev	33.1	203.	3.8
%RSD	.49393	.48693	.18234
#1	6717.5	41536.	2115.0
#2	6668.0	41646.	2110.1
#3	6730.8	41930.	2107.4

Sample Name: L1606109802 Acquired: 6/23/2016 17:42:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00131	.04120	.00085	.00475	.00003	.00006	.12679	.00004
Stddev	.00226	.00310	.00173	.00253	.00122	.00003	.01985	.00014
%RSD	172.44	7.5124	203.60	53.268	4293.6	44.669	15.655	354.39
#1	-.00391	.03975	.00123	.00360	.00116	.00006	.11393	-.00006
#2	-.00005	.04475	-.00104	.00766	.00019	.00008	.11680	.00020
#3	.00004	.03909	.00235	.00301	-.00126	.00003	.14965	-.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00021	.00349	.00237	.18011	.09946	-.00236	.07943	.00699
Stddev	.00033	.00204	.00114	.02015	.07387	.00179	.02978	.00080
%RSD	156.99	58.571	48.052	11.185	74.275	75.957	37.492	11.493
#1	.00015	.00129	.00134	.19792	.08997	-.00029	.05468	.00760
#2	-.00027	.00385	.00359	.18417	.17761	-.00335	.11249	.00608
#3	-.00051	.00533	.00218	.15825	.03079	-.00344	.07113	.00730
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00657	.08383	.00108	-.00632	.00042	.00155	-.00209	-.00610
Stddev	.00027	.01857	.00084	.00291	.00245	.00366	.00646	.00130
%RSD	4.0493	22.151	77.513	46.079	588.34	235.09	308.97	21.310
#1	.00687	.08551	.00034	-.00968	.00101	-.00057	-.00876	-.00625
#2	.00646	.10150	.00091	-.00476	.00251	.00578	.00414	-.00474
#3	.00637	.06448	.00199	-.00452	-.00227	-.00054	-.00165	-.00733
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606109802 Acquired: 6/23/2016 17:42:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00054**	**.00097**	**.00153**	**-.00310**	**-.00091**	**.00301**	**.27773**
Stddev	.00104	.00057	.00430	.00238	.00037	.00011	.05024
%RSD	194.32	58.620	280.21	76.638	40.878	3.7408	18.089
#1	-.00046	.00112	.00644	-.00519	-.00093	.00307	.33422
#2	.00046	.00144	-.00158	-.00051	-.00052	.00288	.23804
#3	-.00161	.00034	-.00025	-.00360	-.00127	.00309	.26094
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6708.6**	**41920.**	**2143.4**
Stddev	42.4	304.	28.2
%RSD	.63177	.72485	1.3166
#1	6687.7	42255.	2139.9
#2	6680.7	41839.	2117.2
#3	6757.4	41664.	2173.3

Sample Name: L1606109802PS Acquired: 6/23/2016 17:46:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573952-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20295	4.9978	.19604	.97111	.51605	.02475	5.3316	.02478
Stddev	.00344	.0037	.00197	.00223	.00051	.00014	.0490	.00015
%RSD	1.6972	.07436	1.0027	.22973	.09917	.56068	.91841	.59158
#1	.19936	4.9935	.19463	.96974	.51663	.02489	5.2786	.02492
#2	.20326	5.0002	.19520	.96990	.51565	.02462	5.3752	.02463
#3	.20622	4.9996	.19828	.97368	.51588	.02474	5.3410	.02480
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10156	.25473	.25812	2.2174	25.577	.52251	5.2379	.25866
Stddev	.00037	.00076	.00168	.0428	.016	.00141	.1063	.00557
%RSD	.36206	.29734	.65016	1.9294	.06131	.27078	2.0291	2.1526
#1	.10143	.25529	.25897	2.1982	25.585	.52373	5.2689	.26416
#2	.10128	.25504	.25619	2.2664	25.559	.52096	5.3253	.25303
#3	.10198	.25387	.25920	2.1875	25.587	.52284	5.1196	.25880
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.51904	25.731	.25472	4.8394	.25146	.60443	.18808	2.5389
Stddev	.00078	.051	.00178	.0161	.00652	.00703	.00446	.0097
%RSD	.15079	.19825	.69859	.33289	2.5939	1.1635	2.3736	.38131
#1	.51985	25.790	.25678	4.8266	.24420	.60484	.18767	2.5333
#2	.51829	25.701	.25365	4.8340	.25681	.59720	.18383	2.5332
#3	.51899	25.702	.25373	4.8575	.25338	.61125	.19273	2.5500
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606109802PS Acquired: 6/23/2016 17:46:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573952-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50319	.52125	.51263	.24941	.49856	.50658	.41681
Stddev	.00067	.00121	.00461	.00337	.00102	.00107	.18065
%RSD	.13272	.23307	.89973	1.3499	.20406	.21178	43.340
#1	.50331	.52232	.50810	.25329	.49739	.50643	.22345
#2	.50379	.52148	.51246	.24723	.49912	.50559	.44572
#3	.50247	.51993	.51732	.24772	.49918	.50772	.58125
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6634.2	41367.	2120.2
Stddev	45.5	136.	31.3
%RSD	.68551	.32832	1.4752
#1	6675.0	41309.	2085.8
#2	6642.5	41522.	2147.0
#3	6585.2	41269.	2127.8

Sample Name: L1606109802SDL Acquired: 6/23/2016 17:49:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573952-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00136	.01502	.00308	.00439	-.00102	.00008	.05957	.00033
Stddev	.00139	.00670	.00298	.00095	.00092	.00003	.03476	.00004
%RSD	102.34	44.577	96.746	21.522	90.289	33.805	58.357	13.382
#1	.00126	.01153	.00023	.00498	-.00197	.00010	.02372	.00038
#2	.00281	.01080	.00283	.00489	-.00093	.00009	.06187	.00030
#3	.00002	.02274	.00619	.00330	-.00015	.00005	.09313	.00031
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00008	-.00017	.00113	.04078	.05810	.00210	.06165	.00226
Stddev	.00020	.00084	.00062	.00773	.11892	.00364	.08929	.00317
%RSD	251.06	498.56	54.543	18.955	204.67	172.87	144.83	140.01
#1	-.00031	.00011	.00044	.04552	-.07722	.00243	-.01365	-.00138
#2	.00002	.00049	.00161	.03186	.14597	-.00169	.03832	.00380
#3	.00006	-.00111	.00136	.04496	.10555	.00557	.16029	.00437
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00155	.03886	.00061	.00728	.00025	-.00084	-.00741	-.00785
Stddev	.00019	.02607	.00103	.00300	.00248	.00463	.00531	.00300
%RSD	12.485	67.084	170.41	41.143	1006.5	550.34	71.645	38.154
#1	.00155	.03807	-.00057	.00580	-.00067	-.00286	-.00354	-.01115
#2	.00174	.01319	.00103	.00532	.00306	.00445	-.01346	-.00530
#3	.00136	.06531	.00136	.01073	-.00164	-.00411	-.00522	-.00710
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606109802SDL Acquired: 6/23/2016 17:49:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573952-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00088**	**.00057**	**.00094**	**-.00409**	**.00019**	**.00149**	**.45954**
Stddev	.00075	.00043	.01378	.00250	.00054	.00013	.21882
%RSD	85.215	74.163	1462.5	61.137	279.61	9.0143	47.616
#1	-.00150	.00022	-.00662	-.00460	.00077	.00151	.33705
#2	-.00005	.00045	.01685	-.00138	-.00029	.00135	.71218
#3	-.00110	.00105	-.00740	-.00631	.00009	.00162	.32941
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6959.1**	**43239.**	**2203.0**
Stddev	42.9	568.	21.3
%RSD	.61702	1.3134	.96538
#1	6931.5	42983.	2192.4
#2	7008.6	43890.	2227.5
#3	6937.3	42844.	2189.1

Sample Name: L1606109802SDL Acquired: 6/23/2016 17:53:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573952-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00025**	**-.00848**	**.00213**	**-.00007**	**-.00079**	**-.00002**	**.00503**
Stddev	.00182	.00079	.00051	.00320	.00067	.00006	.00717
%RSD	715.84	9.3659	23.973	4723.2	84.264	337.91	142.65
#1	.00063	-.00798	.00169	-.00266	-.00014	-.00004	.01305
#2	.00185	-.00807	.00200	-.00107	-.00147	-.00006	.00280
#3	-.00172	-.00940	.00269	.00352	-.00077	.00005	-.00077
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00019**	**.00018**	**.00059**	**-.00006**	**-.00347**	**.12523**	**.00146**
Stddev	.00022	.00029	.00107	.00164	.02019	.11611	.00742
%RSD	115.68	162.83	180.16	2901.7	582.00	92.721	507.00
#1	-.00042	.00014	.00077	.00098	.00022	-.00797	-.00642
#2	-.00018	.00048	-.00055	.00080	-.02525	.20510	.00830
#3	.00002	-.00009	.00157	-.00195	.01463	.17855	.00251
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.04612**	**.00275**	**-.00011**	**.02502**	**-.00021**	**.01583**	**.00003**
Stddev	.09471	.00202	.00043	.02748	.00019	.00290	.00282
%RSD	205.33	73.453	401.99	109.83	87.851	18.294	10124.
#1	.13667	.00063	.00029	.00253	-.00023	.01917	-.00119
#2	-.05226	.00465	-.00004	.01688	-.00002	.01426	.00325
#3	.05396	.00298	-.00056	.05566	-.00039	.01406	-.00198
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606109802SDL Acquired: 6/23/2016 17:53:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573952-02

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00091**	F **-.01196**	**-.01108**	**-.00050**	**.00003**	**-.00725**	**-.00124**
Stddev	.00148	.00527	.00232	.00100	.00102	.00296	.00502
%RSD	162.61	44.027	20.977	201.11	3086.6	40.796	405.85
#1	-.00253	-.00590	-.01377	-.00126	-.00114	-.00905	-.00509
#2	.00038	-.01454	-.00977	-.00086	.00061	-.00384	-.00306
#3	-.00058	-.01544	-.00971	.00063	.00062	-.00886	.00444
Check ?	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit		90.000					
Low Limit		-.01000					

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00077**	**.00084**	**.22312**
Stddev	.00068	.00020	.30532
%RSD	88.979	23.810	136.84
#1	-.00155	.00083	.52434
#2	-.00032	.00104	-.08613
#3	-.00043	.00064	.23116
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6981.3**	**43420.**	**2189.5**
Stddev	46.4	74.	50.4
%RSD	.66474	.17141	2.3033
#1	6946.3	43470.	2150.2
#2	7033.9	43456.	2171.9
#3	6963.6	43335.	2246.4

Sample Name: CCV Acquired: 6/23/2016 17:57:55 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39465	9.8592	.39426	.49051	.99575	.04951	9.9296
Stddev	.00130	.0543	.00077	.00725	.01287	.00025	.0980
%RSD	.33011	.55085	.19471	1.4789	1.2926	.50940	.98742
#1	.39334	9.8260	.39515	.49135	.98121	.04933	9.8278
#2	.39467	9.9219	.39384	.49731	1.0057	.04980	9.9378
#3	.39594	9.8298	.39380	.48287	1.0003	.04940	10.023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04931	.19809	.49689	.49699	3.9877	49.594	1.0016
Stddev	.00035	.00039	.00097	.00140	.0463	.339	.0098
%RSD	.69996	.19569	.19580	.28221	1.1620	.68359	.97475
#1	.04892	.19768	.49580	.49756	3.9856	49.220	.99073
#2	.04957	.19845	.49718	.49801	3.9424	49.882	1.0096
#3	.04943	.19814	.49768	.49539	4.0350	49.679	1.0045
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.8713	.49596	.98783	49.518	.49712	9.8043	.49873
Stddev	.0582	.00613	.00477	.522	.00109	.0177	.00163
%RSD	.58979	1.2355	.48297	1.0550	.21900	.18012	.32630
#1	9.8044	.48970	.99184	48.919	.49799	9.7884	.49800
#2	9.9106	.50194	.98911	49.882	.49748	9.8233	.49759
#3	9.8989	.49623	.98255	49.751	.49590	9.8012	.50059
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 17:57:55 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1826**	**.39764**	**4.9717**	**.98502**	**.99375**	**.98233**	**.49603**
Stddev	.0050	.01088	.0089	.00253	.01310	.00868	.00327
%RSD	.42000	2.7362	.17868	.25689	1.3184	.88336	.65997
#1	1.1879	.39017	4.9692	.98339	.97877	.98140	.49964
#2	1.1780	.41013	4.9816	.98794	1.0031	.99143	.49324
#3	1.1820	.39263	4.9643	.98374	.99940	.97415	.49523
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.99379**	**.99204**	F **.18789**
Stddev	.00563	.00149	.09910
%RSD	.56658	.15028	52.747
#1	.98993	.99220	.30141
#2	1.0003	.99344	.11863
#3	.99118	.99047	.14362
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6631.8**	**40728.**	**2099.0**
Stddev	45.9	503.	16.5
%RSD	.69181	1.2341	.78530
#1	6611.3	41075.	2107.2
#2	6599.8	40151.	2080.0
#3	6684.4	40956.	2109.8

Sample Name: CCB Acquired: 6/23/2016 18:01:33 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00036	.01129	.00078	.00451	-.00045	.00008	-.00465
Stddev	.00148	.00039	.00155	.00327	.00048	.00013	.02449
%RSD	411.84	3.4907	199.15	72.530	105.95	171.62	526.45
#1	-.00053	.01164	.00012	.00090	.00005	.00000	-.02544
#2	.00207	.01087	.00255	.00727	-.00049	-.00000	.02235
#3	-.00046	.01137	-.00034	.00537	-.00091	.00022	-.01087
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00012	.00018	.00054	.00011	-.01088	.09874	.00114
Stddev	.00029	.00030	.00119	.00044	.03748	.07317	.00332
%RSD	246.95	168.18	220.68	381.54	344.38	74.101	290.12
#1	.00045	.00022	-.00081	.00062	.01794	.03451	-.00267
#2	.00000	.00046	.00100	-.00013	-.05325	.08332	.00343
#3	-.00010	-.00014	.00143	-.00015	.00266	.17839	.00267
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00689	.00204	.00366	.02542	-.00029	.00140	-.00069
Stddev	.06132	.00361	.00131	.02117	.00182	.00788	.00343
%RSD	889.59	177.04	35.712	83.279	629.54	563.60	498.72
#1	.06069	-.00150	.00240	.01760	.00160	.00084	.00259
#2	-.02239	.00571	.00357	.00928	-.00043	.00955	-.00039
#3	-.05898	.00190	.00501	.04939	-.00204	-.00619	-.00426
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 18:01:33 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00612**	**-.00027**	**.00319**	**.00025**	**.00023**	**-.00325**	**-.00471**
Stddev	.00153	.00342	.00271	.00103	.00038	.00279	.00086
%RSD	25.064	1269.5	84.974	412.62	169.30	85.836	18.269
#1	.00449	-.00355	.00269	-.00074	-.00018	-.00174	-.00566
#2	.00753	-.00053	.00612	.00132	.00028	-.00154	-.00398
#3	.00634	.00327	.00077	.00017	.00058	-.00647	-.00449
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.00008**	**.00027**	F **.24105**
Stddev	.00092	.00012	.41150
%RSD	1209.9	46.847	170.71
#1	-.00098	.00021	.54842
#2	.00054	.00018	.40117
#3	.00067	.00041	-.22644
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6672.3**	**41768.**	**2119.7**
Stddev	57.6	262.	32.6
%RSD	.86323	.62780	1.5373
#1	6705.6	42016.	2128.0
#2	6705.5	41794.	2083.8
#3	6605.8	41493.	2147.4

Sample Name: L1606109803 Acquired: 6/23/2016 18:05:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00032**	**.03085**	**-.00127**	**.00183**	**-.00146**	**.00004**	**.19954**	**.00007**
Stddev	.00236	.01045	.00348	.00087	.00085	.00007	.02976	.00016
%RSD	746.66	33.859	274.17	47.779	57.997	158.69	14.912	230.00
#1	.00241	.03757	-.00520	.00147	-.00081	.00010	.20422	-.00011
#2	-.00171	.03617	.00140	.00283	-.00242	.00007	.22668	.00018
#3	-.00165	.01882	-.00001	.00120	-.00116	-.00003	.16772	.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00003**	**.00568**	**.00388**	**.39187**	**.10478**	**.00034**	**.12477**	**.01117**
Stddev	.00033	.00040	.00075	.05326	.14846	.00690	.09992	.00202
%RSD	1223.9	7.0496	19.420	13.590	141.68	2058.5	80.089	18.078
#1	-.00037	.00598	.00367	.36606	.20874	.00375	.01013	.01079
#2	.00028	.00523	.00472	.45311	.17085	.00486	.17078	.00937
#3	.00001	.00584	.00326	.35643	-.06524	-.00761	.19339	.01336
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.01433**	**.14788**	**.00239**	**-.00401**	**-.00121**	**-.00112**	**-.00779**	**.00373**
Stddev	.00040	.04259	.00040	.00288	.00159	.00565	.00213	.00124
%RSD	2.7886	28.802	16.639	71.900	130.64	504.39	27.319	33.311
#1	.01390	.18295	.00247	-.00583	-.00301	.00539	-.01003	.00233
#2	.01440	.16021	.00274	-.00069	-.00001	-.00484	-.00757	.00419
#3	.01469	.10048	.00196	-.00551	-.00062	-.00391	-.00579	.00468
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606109803 Acquired: 6/23/2016 18:05:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00017	.00160	.00432	-.00693	-.00031	.00470	.05902
Stddev	.00059	.00022	.00543	.00465	.00124	.00007	.04433
%RSD	337.40	13.821	125.73	67.076	399.03	1.4400	75.109
#1	.00054	.00138	-.00146	-.01029	-.00033	.00477	.10269
#2	-.00050	.00182	.00932	-.00163	.00094	.00464	.06031
#3	.00048	.00160	.00510	-.00886	-.00154	.00469	.01406
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6701.9	41737.	2133.3
Stddev	17.7	223.	23.6
%RSD	.26482	.53370	1.1066
#1	6684.0	41936.	2107.7
#2	6702.4	41778.	2138.2
#3	6719.4	41496.	2154.1

Sample Name: L1606109804 Acquired: 6/23/2016 18:09:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00003	.00550	.00148	.00398	.00014	-.00000	.05639	.00011
Stddev	.00219	.00612	.00302	.00090	.00066	.00014	.02441	.00010
%RSD	8054.2	111.32	204.56	22.756	478.27	12248.	43.295	91.576
#1	.00205	-.00072	.00110	.00492	.00048	.00000	.08169	.00011
#2	.00034	.00570	.00467	.00389	-.00062	-.00014	.05452	.00020
#3	-.00231	.01151	-.00134	.00312	.00056	.00014	.03297	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	-.00048	.00039	.00509	.07226	-.00432	.08860	-.00098
Stddev	.00047	.00177	.00047	.00413	.11460	.00188	.02282	.00192
%RSD	981.46	369.62	120.47	81.176	158.60	43.602	25.750	195.61
#1	.00046	-.00226	.00002	.00662	.11487	-.00448	.08077	-.00230
#2	.00015	.00128	.00093	.00041	-.05755	-.00612	.11430	.00122
#3	-.00046	-.00045	.00024	.00823	.15944	-.00236	.07074	-.00186
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00038	.28447	-.00051	-.01137	-.00379	.00134	-.00025	-.01961
Stddev	.00030	.03222	.00045	.00801	.00241	.00677	.00472	.00119
%RSD	78.960	11.326	89.242	70.503	63.646	504.08	1916.6	6.0610
#1	.00007	.28282	-.00042	-.00434	-.00635	.00646	.00471	-.01950
#2	.00041	.31748	-.00099	-.02009	-.00156	-.00633	-.00469	-.01848
#3	.00066	.25310	-.00010	-.00967	-.00346	.00390	-.00076	-.02085
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606109804 Acquired: 6/23/2016 18:09:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00008**	**.00727**	**-.00271**	**-.00425**	**.00021**	**.00148**	**-.01678**
Stddev	.00012	.00034	.00158	.00185	.00078	.00021	.29091
%RSD	149.43	4.6162	58.348	43.577	374.98	14.359	1733.8
#1	.00006	.00757	-.00433	-.00448	.00104	.00139	.18184
#2	-.00016	.00734	-.00263	-.00229	.00011	.00173	.11852
#3	-.00014	.00691	-.00117	-.00598	-.00052	.00134	-.35070
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6688.2**	**41655.**	**2109.0**
Stddev	56.4	133.	18.7
%RSD	.84272	.31956	.88651
#1	6659.9	41774.	2087.4
#2	6753.1	41679.	2119.9
#3	6651.7	41511.	2119.7

Sample Name: L1606076401 Acquired: 6/23/2016 18:13:29 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00245	.67019	.00010	.00833	.01433	.00003	37.313	.00009
Stddev	.00064	.00585	.00478	.00400	.00127	.00014	.132	.00011
%RSD	26.141	.87236	4915.8	47.991	8.8257	429.53	.35301	115.20
#1	.00243	.67675	-.00530	.00372	.01377	.00008	37.462	.00013
#2	.00182	.66553	.00380	.01082	.01578	.00015	37.267	-.00003
#3	.00310	.66828	.00179	.01046	.01345	-.00013	37.211	.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00177	.13494	.00493	2.1093	3.5116	.00538	12.516	.47690
Stddev	.00027	.00081	.00173	.0368	.1337	.00492	.251	.00425
%RSD	15.254	.60366	35.121	1.7452	3.8081	91.565	2.0071	.89198
#1	.00146	.13588	.00619	2.1427	3.6345	.01007	12.781	.48122
#2	.00191	.13437	.00564	2.0698	3.5309	.00580	12.487	.47677
#3	.00195	.13459	.00296	2.1153	3.3692	.00026	12.281	.47272
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00239	7.1134	.06661	.12519	.00353	.00172	.00189	5.8796
Stddev	.00028	.0421	.00053	.00389	.00278	.00471	.00179	.0247
%RSD	11.781	.59130	.79827	3.1089	78.703	273.59	94.695	.41985
#1	.00270	7.1321	.06712	.12451	.00584	-.00218	-.00008	5.8957
#2	.00230	7.0652	.06666	.12938	.00431	.00694	.00342	5.8512
#3	.00216	7.1429	.06605	.12168	.00045	.00040	.00234	5.8918
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606076401 Acquired: 6/23/2016 18:13:29 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00001	.21848	.02711	-.00526	.00152	.01133	.29378
Stddev	.00053	.00139	.00262	.00115	.00075	.00027	.28967
%RSD	4257.4	.63577	9.6647	21.837	49.639	2.3502	98.600
#1	.00059	.22007	.03006	-.00608	.00213	.01140	.10070
#2	-.00044	.21748	.02623	-.00576	.00176	.01103	.15379
#3	-.00012	.21789	.02504	-.00395	.00068	.01155	.62685
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6666.0	41778.	2147.7
Stddev	21.7	229.	44.3
%RSD	.32537	.54913	2.0641
#1	6658.2	41696.	2121.5
#2	6690.5	41600.	2122.8
#3	6649.4	42037.	2198.9

Sample Name: L1606076402 Acquired: 6/23/2016 18:17:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00113	.17691	.00122	.01438	.01126	-.00007	47.643	.00046
Stddev	.00008	.00451	.00550	.00105	.00043	.00007	.152	.00035
%RSD	6.6984	2.5493	451.48	7.3081	3.8067	88.547	.31914	77.689
#1	-.00121	.17594	.00265	.01351	.01094	-.00002	47.529	.00010
#2	-.00113	.18183	-.00485	.01554	.01110	-.00005	47.816	.00045
#3	-.00106	.17297	.00586	.01408	.01175	-.00015	47.585	.00081
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00047	.00770	.00173	.25827	3.3547	.00082	13.331	.01953
Stddev	.00066	.00082	.00180	.01292	.0490	.00546	.106	.00183
%RSD	140.89	10.586	103.87	5.0018	1.4609	664.19	.79675	9.3761
#1	-.00028	.00820	-.00034	.25716	3.3133	.00131	13.289	.02164
#2	.00073	.00815	.00288	.24595	3.4088	.00602	13.252	.01835
#3	.00096	.00676	.00265	.27171	3.3420	-.00487	13.451	.01861
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	85.841	.00393	.01987	.00031	.00407	-.00554	4.9290
Stddev	.00036	.284	.00086	.00461	.00230	.00232	.00977	.0097
%RSD	74.124	.33036	21.834	23.192	751.95	56.913	176.51	.19731
#1	.00064	85.671	.00313	.02519	-.00217	.00145	-.01667	4.9400
#2	.00074	85.684	.00381	.01712	.00238	.00583	.00165	4.9218
#3	.00007	86.169	.00484	.01729	.00072	.00495	-.00160	4.9251
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606076402 Acquired: 6/23/2016 18:17:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00172	.17262	-.00265	-.00410	-.00007	.00471	.40901
Stddev	.00082	.00082	.00405	.00250	.00091	.00010	.33393
%RSD	47.322	.47585	153.10	60.989	1359.2	2.0262	81.643
#1	.00252	.17344	-.00309	-.00186	-.00069	.00470	.53676
#2	.00089	.17264	.00161	-.00680	.00097	.00462	.03007
#3	.00176	.17180	-.00646	-.00364	-.00048	.00481	.66021
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6546.5	40659.	2108.5
Stddev	28.1	201.	15.3
%RSD	.42958	.49347	.72557
#1	6561.7	40449.	2124.6
#2	6514.1	40849.	2106.9
#3	6563.7	40679.	2094.1

Sample Name: L1606076403 Acquired: 6/23/2016 18:21:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00137	.15342	.00118	.01561	.01121	.00003	46.708	.00034
Stddev	.00144	.01220	.00354	.00156	.00114	.00007	.333	.00016
%RSD	104.86	7.9497	299.98	10.017	10.172	207.11	.71252	47.105
#1	.00276	.15228	.00523	.01638	.01119	.00004	47.070	.00019
#2	.00148	.16615	-.00041	.01663	.01236	.00010	46.637	.00033
#3	-.00012	.14184	-.00128	.01381	.01008	-.00004	46.416	.00051
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00045	.00847	.00160	.25037	3.2322	-.00343	13.109	.01732
Stddev	.00048	.00215	.00051	.01608	.1197	.00219	.140	.00025
%RSD	105.62	25.328	31.641	6.4231	3.7025	63.908	1.0676	1.4475
#1	.00099	.01061	.00135	.24721	3.1397	-.00591	13.252	.01732
#2	.00010	.00848	.00126	.26779	3.1895	-.00259	13.105	.01706
#3	.00026	.00632	.00218	.23609	3.3674	-.00178	12.972	.01756
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	83.906	.00432	.01477	.00450	.00054	-.00630	4.8394
Stddev	.00021	.704	.00062	.00496	.00594	.00823	.00533	.0108
%RSD	145.51	.83873	14.339	33.621	132.00	1523.4	84.657	.22390
#1	.00024	84.668	.00370	.00922	.00040	-.00564	-.00982	4.8447
#2	.00028	83.772	.00494	.01629	.01131	.00989	-.00890	4.8466
#3	-.00010	83.280	.00432	.01879	.00179	-.00263	-.00016	4.8270
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606076403 Acquired: 6/23/2016 18:21:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00189	.16928	-.00280	-.00350	-.00037	.00427	.33488
Stddev	.00087	.00110	.00261	.00165	.00108	.00027	.28811
%RSD	46.182	.64857	93.235	47.233	289.03	6.2712	86.033
#1	.00253	.17053	-.00577	-.00531	-.00161	.00439	.48732
#2	.00090	.16886	-.00085	-.00207	.00037	.00397	.00258
#3	.00224	.16845	-.00178	-.00311	.00012	.00447	.51476
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6581.0	40913.	2128.6
Stddev	5.4	301.	32.8
%RSD	.08278	.73656	1.5394
#1	6574.8	40952.	2106.9
#2	6584.5	40594.	2112.5
#3	6583.9	41193.	2166.3

Sample Name: L1606095701 Acquired: 6/23/2016 18:25:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00041	.46956	.00251	.02030	.04833	.00003	49.298	.00021
Stddev	.00116	.00247	.00083	.00483	.00100	.00005	.380	.00019
%RSD	283.11	.52507	32.898	23.794	2.0760	143.19	.77063	90.357
#1	.00160	.46729	.00340	.01479	.04946	-.00001	49.398	.00006
#2	-.00072	.47218	.00235	.02380	.04755	.00003	49.618	.00041
#3	.00036	.46920	.00178	.02230	.04798	.00008	48.878	.00014
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	.00117	.00294	.90286	1.5234	.00299	12.797	.07921
Stddev	.00016	.00017	.00150	.02505	.1258	.00416	.204	.00280
%RSD	34.195	14.848	50.801	2.7744	8.2581	138.79	1.5973	3.5404
#1	.00039	.00098	.00204	.92120	1.4450	-.00175	12.561	.07627
#2	.00038	.00131	.00212	.91306	1.4566	.00596	12.905	.08186
#3	.00067	.00123	.00467	.87432	1.6685	.00477	12.925	.07952
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00045	3.8733	.00033	.04837	.00336	-.00283	.00062	4.1822
Stddev	.00048	.0248	.00057	.00397	.00588	.00529	.00509	.0038
%RSD	106.40	.63975	174.89	8.1973	175.22	186.90	821.76	.09190
#1	.00063	3.9009	.00047	.05197	.00822	.00326	-.00493	4.1833
#2	-.00009	3.8529	.00082	.04903	-.00318	-.00552	.00170	4.1779
#3	.00081	3.8662	-.00030	.04412	.00504	-.00623	.00508	4.1853
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606095701 Acquired: 6/23/2016 18:25:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00007	.26575	-.00460	-.00190	.00016	.00477	.18671
Stddev	.00076	.00141	.00397	.00470	.00026	.00014	.25228
%RSD	1127.9	.53205	86.124	248.06	165.00	2.9160	135.12
#1	.00070	.26627	-.00436	-.00420	.00020	.00486	-.03830
#2	-.00009	.26684	-.00077	.00352	.00040	.00485	.13898
#3	-.00081	.26416	-.00869	-.00501	-.00012	.00461	.45944
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6633.8	41376.	2099.5
Stddev	37.1	291.	7.1
%RSD	.55940	.70297	.33861
#1	6661.1	41062.	2100.6
#2	6648.8	41636.	2091.9
#3	6591.5	41431.	2106.0

Sample Name: L1606095702 Acquired: 6/23/2016 18:29:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00036	.02149	-.00029	.01798	.04394	-.00005	48.548	.00016
Stddev	.00120	.00717	.00363	.00283	.00013	.00009	.113	.00020
%RSD	338.55	33.343	1265.1	15.710	.28482	186.99	.23191	128.29
#1	.00155	.02937	-.00024	.01478	.04391	-.00003	48.543	-.00006
#2	.00038	.01536	-.00394	.01905	.04383	.00003	48.664	.00019
#3	-.00086	.01975	.00332	.02012	.04408	-.00015	48.439	.00034
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00023	.00018	.00270	.01333	1.4334	.00181	12.360	.04977
Stddev	.00029	.00074	.00104	.00520	.0107	.00516	.188	.00147
%RSD	124.55	409.37	38.421	39.005	.74428	284.70	1.5241	2.9581
#1	.00010	-.00042	.00376	.01807	1.4222	-.00050	12.567	.04817
#2	-.00039	.00100	.00266	.00777	1.4347	.00772	12.199	.05008
#3	-.00041	-.00004	.00169	.01416	1.4434	-.00179	12.314	.05107
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00046	3.8295	.00047	.00890	-.00152	.00413	-.00556	3.5646
Stddev	.00041	.0622	.00056	.00959	.00091	.00142	.00430	.0105
%RSD	89.215	1.6239	120.78	107.82	59.749	34.322	77.442	.29519
#1	.00025	3.8845	.00007	.00556	-.00178	.00386	-.00914	3.5766
#2	.00020	3.7620	.00111	.00142	-.00226	.00566	-.00674	3.5599
#3	.00094	3.8419	.00022	.01971	-.00051	.00286	-.00079	3.5572
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606095702 Acquired: 6/23/2016 18:29:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00032	.25718	-.01617	-.00292	-.00158	.00276	.25728
Stddev	.00011	.00214	.00935	.00156	.00059	.00016	.40159
%RSD	34.052	.83289	57.848	53.296	37.039	5.8090	156.09
#1	.00022	.25955	-.02691	-.00442	-.00103	.00282	-.15534
#2	.00044	.25539	-.00976	-.00131	-.00219	.00288	.64684
#3	.00029	.25660	-.01185	-.00303	-.00152	.00258	.28035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6634.1	41520.	2136.4
Stddev	35.2	131.	44.7
%RSD	.53089	.31504	2.0909
#1	6598.3	41376.	2086.8
#2	6635.3	41551.	2148.8
#3	6668.7	41632.	2173.6

Sample Name: CCV Acquired: 6/23/2016 18:33:11 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.37892	9.4387	.37738	.47256	.94946	.04758	9.5075
Stddev	.00148	.0146	.00259	.00305	.00322	.00017	.0685
%RSD	.39011	.15488	.68700	.64544	.33864	.35065	.72027
#1	.38009	9.4240	.38032	.47530	.94608	.04771	9.4676
#2	.37942	9.4533	.37541	.46928	.94982	.04765	9.4684
#3	.37726	9.4389	.37641	.47312	.95247	.04739	9.5866
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04694	.19097	.47642	.48024	3.7454	47.589	.96391
Stddev	.00038	.00060	.00177	.00172	.0267	.209	.00531
%RSD	.81436	.31192	.37218	.35728	.71307	.43924	.55081
#1	.04714	.19166	.47496	.48079	3.7624	47.380	.95931
#2	.04717	.19062	.47589	.48161	3.7146	47.589	.96972
#3	.04649	.19064	.47839	.47831	3.7591	47.798	.96271
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.5330	.47716	.95180	47.369	.47779	9.4288	.47264
Stddev	.1380	.00452	.00445	.136	.00101	.0198	.00044
%RSD	1.4480	.94690	.46703	.28759	.21184	.20944	.09281
#1	9.3868	.47739	.95630	47.213	.47768	9.4450	.47285
#2	9.6611	.47253	.95168	47.430	.47683	9.4068	.47292
#3	9.5511	.48156	.94742	47.463	.47885	9.4347	.47213
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 18:33:11 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1353**	**.38157**	**4.7878**	**.94812**	**.94630**	**.94801**	**.48121**
Stddev	.0015	.00111	.0073	.00312	.00312	.01023	.00217
%RSD	.12990	.29010	.15352	.32951	.33000	1.0787	.45189
#1	1.1366	.38222	4.7963	.95070	.94271	.93631	.48083
#2	1.1337	.38220	4.7828	.94465	.94841	.95529	.47925
#3	1.1356	.38030	4.7843	.94900	.94778	.95241	.48355
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.95198**	**.95325**	F **.09466**
Stddev	.00202	.00225	.27239
%RSD	.21241	.23553	287.76
#1	.95399	.95580	.19384
#2	.95199	.95239	-.21343
#3	.94995	.95157	.30356
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6829.4**	**42429.**	**2191.9**
Stddev	12.0	491.	25.5
%RSD	.17610	1.1575	1.1649
#1	6837.9	42290.	2175.1
#2	6834.7	42022.	2221.3
#3	6815.6	42975.	2179.4

Sample Name: CCB Acquired: 6/23/2016 18:36:49 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00322**	**.00107**	**-.00054**	**.00322**	**.00006**	**.00009**	**.02193**
Stddev	.00043	.00584	.00138	.00071	.00147	.00001	.03302
%RSD	13.435	544.37	257.70	22.014	2255.3	10.149	150.58
#1	.00309	.00723	.00106	.00367	-.00132	.00008	.01644
#2	.00371	.00039	-.00144	.00240	.00160	.00009	-.00800
#3	.00287	-.00440	-.00123	.00358	-.00009	.00010	.05735
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**.00018**	**-.00042**	**.00126**	**.02179**	**.11279**	**-.00067**
Stddev	.00014	.00038	.00167	.00134	.02326	.08636	.00587
%RSD	182.22	210.03	399.06	106.42	106.75	76.563	872.21
#1	.00024	.00056	-.00036	.00176	.03802	.20919	-.00743
#2	.00004	-.00020	.00122	-.00026	-.00486	.08667	.00222
#3	-.00004	.00019	-.00212	.00226	.03219	.04251	.00319
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02601**	**.00041**	**.00309**	**.08611**	**-.00017**	**.00201**	**-.00231**
Stddev	.00787	.00462	.00037	.06126	.00074	.00369	.00149
%RSD	30.276	1132.0	12.029	71.145	441.11	183.39	64.604
#1	.02246	-.00052	.00267	.06682	.00059	-.00054	-.00062
#2	.02053	.00542	.00337	.03682	-.00020	.00033	-.00285
#3	.03503	-.00368	.00324	.15470	-.00090	.00625	-.00346
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 18:36:49 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00178**	**.00235**	**.00300**	**.00030**	**.00073**	**-.00027**	**-.00344**
Stddev	.00352	.00736	.00017	.00075	.00010	.00192	.00271
%RSD	198.25	313.53	5.6604	253.34	13.143	700.32	78.666
#1	-.00221	.00539	.00284	.00092	.00062	-.00182	-.00634
#2	.00305	-.00605	.00318	-.00054	.00077	.00188	-.00302
#3	.00448	.00771	.00299	.00050	.00080	-.00088	-.00097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.00023**	**.00017**	F **.30996**
Stddev	.00082	.00026	.44738
%RSD	353.90	152.67	144.34
#1	-.00013	.00031	.51733
#2	.00118	-.00013	-.20349
#3	-.00035	.00032	.61602
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6847.7**	**42764.**	**2190.8**
Stddev	18.6	212.	5.5
%RSD	.27202	.49617	.24955
#1	6838.4	42826.	2184.7
#2	6869.1	42938.	2192.8
#3	6835.5	42527.	2195.1

Sample Name: L1606095703 Acquired: 6/23/2016 18:40:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00157	.59442	.00345	.01544	.05978	.00002	36.179	.00021
Stddev	.00140	.00140	.00372	.00226	.00166	.00008	.233	.00017
%RSD	89.489	.23471	107.82	14.612	2.7840	315.39	.64331	79.288
#1	.00259	.59283	.00035	.01648	.06092	.00011	36.405	.00023
#2	.00214	.59501	.00758	.01285	.06056	-.00003	36.192	.00038
#3	-.00003	.59543	.00243	.01699	.05787	-.00000	35.940	.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00050	.00171	.00235	.99066	1.0189	.00306	4.2466	.34135
Stddev	.00029	.00065	.00116	.00191	.1477	.00520	.1543	.00578
%RSD	57.273	38.095	49.421	.19284	14.493	170.05	3.6338	1.6936
#1	.00018	.00100	.00118	.99000	.97647	.00895	4.1429	.34787
#2	.00074	.00184	.00237	.98917	1.1832	.00113	4.4240	.33684
#3	.00059	.00229	.00350	.99281	.89714	-.00090	4.1730	.33935
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00158	9.9139	.00125	.06880	.00203	.00079	-.00517	4.4030
Stddev	.00038	.0706	.00076	.00768	.00168	.00577	.00107	.0083
%RSD	24.292	.71237	60.961	11.168	82.575	735.03	20.696	.18913
#1	.00191	9.9842	.00113	.06853	.00010	.00362	-.00400	4.4025
#2	.00167	9.8430	.00208	.07661	.00292	.00459	-.00539	4.3950
#3	.00116	9.9146	.00056	.06125	.00308	-.00586	-.00611	4.4116
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606095703 Acquired: 6/23/2016 18:40:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00017	.13708	.00451	-.00202	.00078	.00490	.22494
Stddev	.00088	.00079	.01428	.00267	.00035	.00024	.11125
%RSD	513.50	.57511	316.39	132.43	44.919	4.8544	49.458
#1	.00022	.13772	-.01170	-.00500	.00067	.00502	.12100
#2	.00102	.13733	.01523	-.00120	.00118	.00506	.21152
#3	-.00073	.13620	.01000	.00015	.00050	.00463	.34229
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6641.1	41542.	2110.2
Stddev	47.0	387.	42.7
%RSD	.70732	.93262	2.0250
#1	6610.2	41137.	2074.7
#2	6618.0	41909.	2098.4
#3	6695.1	41580.	2157.7

Sample Name: L1606095704 Acquired: 6/23/2016 18:44:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00171	.01427	-.00254	.01401	.04366	.00001	35.183	.00015
Stddev	.00187	.00827	.00245	.00145	.00176	.00007	.524	.00017
%RSD	109.40	57.926	96.490	10.332	4.0281	938.95	1.4883	116.97
#1	.00375	.01991	-.00531	.01561	.04341	-.00008	35.381	.00027
#2	.00007	.00478	-.00065	.01361	.04552	.00004	35.579	-.00005
#3	.00132	.01813	-.00167	.01280	.04203	.00006	34.589	.00022
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00027	.00023	.00314	.00854	.90554	-.00053	4.1738	.13131
Stddev	.00017	.00027	.00209	.01235	.04311	.00195	.0469	.00168
%RSD	61.984	120.89	66.548	144.50	4.7611	368.87	1.1239	1.2783
#1	-.00041	.00004	.00554	-.00223	.92277	.00023	4.1196	.13296
#2	-.00032	.00054	.00170	.02202	.93737	.00093	4.2002	.13137
#3	-.00008	.00010	.00219	.00585	.85648	-.00275	4.2015	.12960
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00146	9.9224	.00122	.02546	.00187	.00274	-.00609	3.6801
Stddev	.00034	.2004	.00114	.01084	.00327	.00376	.00550	.0062
%RSD	23.395	2.0192	93.082	42.580	174.85	137.20	90.310	.16848
#1	.00107	10.130	.00251	.01415	.00529	.00275	-.00747	3.6873
#2	.00171	9.9074	.00082	.02646	.00154	-.00102	-.01077	3.6761
#3	.00161	9.7299	.00034	.03577	-.00122	.00649	-.00003	3.6770
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606095704 Acquired: 6/23/2016 18:44:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00044**	**.13398**	**-.00259**	**-.00487**	**-.00042**	**.00404**	**.39599**
Stddev	.00030	.00097	.00767	.00290	.00014	.00020	.04470
%RSD	68.069	.72204	296.19	59.450	32.889	5.0229	11.288
#1	-.00031	.13510	.00380	-.00161	-.00033	.00423	.39112
#2	-.00079	.13343	-.00047	-.00585	-.00058	.00406	.35393
#3	-.00024	.13342	-.01109	-.00716	-.00035	.00383	.44293
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6691.6**	**41683.**	**2130.5**
Stddev	16.3	188.	44.8
%RSD	.24349	.45009	2.1019
#1	6674.7	41874.	2106.4
#2	6707.2	41499.	2102.9
#3	6692.9	41675.	2182.2

Sample Name: L1606095705 Acquired: 6/23/2016 18:48:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00080	.18137	.00132	.01393	.05227	.00000	47.576	-.00017
Stddev	.00022	.00236	.00336	.00262	.00068	.00006	.406	.00016
%RSD	27.412	1.3014	254.59	18.802	1.2936	1545.4	.85245	98.177
#1	.00055	.17984	-.00165	.01122	.05295	.00007	47.676	-.00014
#2	.00093	.18018	.00064	.01413	.05226	-.00001	47.129	-.00002
#3	.00093	.18409	.00497	.01644	.05160	-.00005	47.922	-.00034
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00015	.00081	.00165	.24818	.66421	-.00237	5.8917	.04016
Stddev	.00049	.00111	.00028	.02329	.15036	.00580	.1105	.00146
%RSD	327.66	136.79	16.738	9.3859	22.638	244.94	1.8758	3.6247
#1	.00050	.00005	.00195	.26948	.57304	-.00408	6.0110	.04145
#2	.00036	.00030	.00162	.22331	.58183	.00409	5.7928	.04046
#3	-.00041	.00208	.00140	.25176	.83776	-.00711	5.8714	.03858
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00054	4.2444	.00044	.02145	-.00147	.00210	-.00161	3.9144
Stddev	.00037	.0382	.00138	.00960	.00364	.00064	.00655	.0080
%RSD	69.362	.90092	314.47	44.765	248.05	30.375	406.75	.20417
#1	.00089	4.2867	-.00079	.03237	-.00565	.00257	-.00833	3.9237
#2	.00058	4.2341	.00193	.01434	.00101	.00137	-.00126	3.9100
#3	.00015	4.2123	.00017	.01764	.00024	.00235	.00476	3.9097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606095705 Acquired: 6/23/2016 18:48:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00124**	**.16651**	**-.01359**	**-.00521**	**.00084**	**.00210**	**.15995**
Stddev	.00099	.00174	.00564	.00062	.00055	.00006	.33370
%RSD	79.444	1.0457	41.504	11.980	65.519	2.7028	208.63
#1	-.00013	.16843	-.01365	-.00569	.00109	.00214	.50631
#2	-.00159	.16606	-.00792	-.00450	.00021	.00203	.13299
#3	-.00201	.16504	-.01920	-.00543	.00123	.00212	-.15945
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6676.1**	**41561.**	**2103.6**
Stddev	12.6	161.	24.0
%RSD	.18881	.38711	1.1405
#1	6690.1	41742.	2107.2
#2	6672.8	41433.	2125.6
#3	6665.6	41508.	2078.0

Sample Name: L1606096101 Acquired: 6/23/2016 18:52:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00082	.01430	.00369	.01759	.00065	-.00004	.04649	-.00016
Stddev	.00101	.00586	.00264	.00153	.00060	.00005	.04475	.00011
%RSD	122.72	40.990	71.487	8.6840	91.837	113.85	96.246	67.330
#1	-.00034	.00813	.00410	.01814	.00134	-.00010	.05237	-.00005
#2	.00136	.01498	.00087	.01587	.00033	.00000	.08801	-.00016
#3	.00144	.01979	.00609	.01877	.00028	-.00003	-.00090	-.00028
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00034	.00024	.01766	.09903	.10656	.00252	-.01866	.00398
Stddev	.00056	.00013	.00069	.02628	.09276	.00342	.06339	.00246
%RSD	164.44	55.453	3.9062	26.535	87.051	136.01	339.78	61.709
#1	.00024	.00020	.01732	.09611	.01800	.00645	-.08612	.00647
#2	-.00016	.00038	.01845	.12665	.09866	.00026	.03967	.00155
#3	.00095	.00013	.01720	.07434	.20302	.00083	-.00952	.00393
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00003	107.16	-.00002	.03665	.00304	.00296	-.00309	5.9627
Stddev	.00055	1.21	.00170	.00369	.00053	.00129	.00524	.0081
%RSD	1743.4	1.1249	7247.6	10.064	17.513	43.498	169.55	.13655
#1	.00064	108.50	.00194	.03929	.00277	.00155	-.00074	5.9648
#2	-.00012	106.16	-.00110	.03243	.00269	.00407	.00057	5.9537
#3	-.00042	106.82	-.00091	.03823	.00365	.00327	-.00910	5.9696
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606096101 Acquired: 6/23/2016 18:52:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00123**	**.00009**	**-.00012**	**-.00673**	**-.00085**	**.00397**	**.21821**
Stddev	.00042	.00025	.00564	.00413	.00089	.00018	.04757
%RSD	34.134	265.03	4817.0	61.413	103.95	4.5901	21.800
#1	.00144	-.00013	-.00483	-.00336	-.00072	.00394	.16868
#2	.00075	.00004	.00613	-.01134	-.00180	.00381	.26354
#3	.00150	.00037	-.00165	-.00548	-.00004	.00417	.22241
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6592.8**	**41252.**	**2103.4**
Stddev	17.0	291.	47.5
%RSD	.25725	.70443	2.2575
#1	6595.4	40916.	2058.8
#2	6574.6	41415.	2153.3
#3	6608.2	41424.	2098.1

Sample Name: L1606096102 Acquired: 6/23/2016 18:56:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00079**	**.03478**	**.00589**	**.01533**	**.02658**	**-.00009**	**27.946**	**-.00005**
Stddev	.00186	.00579	.00137	.00335	.00018	.00004	.659	.00022
%RSD	235.75	16.640	23.210	21.834	.65976	40.468	2.3584	471.56
#1	.00069	.02830	.00447	.01574	.02649	-.00010	28.691	-.00000
#2	-.00287	.03661	.00719	.01180	.02646	-.00005	27.440	-.00029
#3	-.00019	.03942	.00601	.01846	.02678	-.00013	27.706	.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00010**	**.00057**	**.00211**	**.07517**	**.95068**	**-.00017**	**4.4154**	**.12814**
Stddev	.00012	.00116	.00118	.01917	.12455	.00106	.0966	.00422
%RSD	122.50	204.26	55.846	25.503	13.101	623.39	2.1883	3.2924
#1	.00017	.00092	.00222	.09509	.80692	-.00137	4.5065	.13251
#2	.00017	.00152	.00323	.05686	1.0194	.00064	4.3140	.12782
#3	-.00004	-.00073	.00088	.07355	1.0258	.00022	4.4257	.12409
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00077**	**2.8960**	**.00155**	**.00675**	**-.00147**	**-.00079**	**-.00450**	**1.9147**
Stddev	.00018	.0986	.00062	.00402	.00109	.00203	.00538	.0058
%RSD	22.925	3.4051	39.619	59.594	74.555	256.57	119.66	.30207
#1	.00069	2.9809	.00179	.00220	-.00127	.00048	-.01064	1.9214
#2	.00097	2.7878	.00202	.00820	-.00048	-.00314	-.00059	1.9118
#3	.00065	2.9192	.00086	.00985	-.00264	.00028	-.00227	1.9110
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606096102 Acquired: 6/23/2016 18:56:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00047**	**.09203**	**-.00699**	**-.00581**	**.00001**	**.00122**	**.19167**
Stddev	.00152	.00238	.00325	.00218	.00060	.00015	.39664
%RSD	319.98	2.5851	46.457	37.509	4349.6	12.672	206.94
#1	.00001	.09410	-.00457	-.00773	.00065	.00107	.60420
#2	-.00217	.08943	-.01068	-.00344	-.00006	.00121	.15771
#3	.00074	.09257	-.00572	-.00625	-.00055	.00138	-.18689
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6690.9**	**41659.**	**2159.3**
Stddev	15.4	166.	23.9
%RSD	.22972	.39942	1.1048
#1	6679.1	41827.	2132.7
#2	6708.3	41655.	2166.7
#3	6685.3	41494.	2178.6

Sample Name: L1606099102 Acquired: 6/23/2016 19:00:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.18278	.01279	.07089	.02678	.00003	1.1753
Stddev	.00017	.00874	.00230	.00310	.00094	.00004	.0130
%RSD	197.98	4.7790	18.004	4.3721	3.5015	169.52	1.1019
#1	.00028	.18746	.01165	.06786	.02601	.00006	1.1606
#2	.00005	.18819	.01543	.07075	.02651	-.00002	1.1806
#3	-.00007	.17271	.01127	.07405	.02782	.00005	1.1848
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00374	.00173	.10547	.01142	.70686	2.4306	.00285
Stddev	.00031	.00017	.00069	.00091	.00672	.2276	.00410
%RSD	8.3801	9.8737	.65029	7.9351	.95134	9.3634	143.93
#1	.00362	.00160	.10540	.01211	.70453	2.2038	.00005
#2	.00409	.00192	.10482	.01039	.71445	2.4290	.00095
#3	.00349	.00167	.10619	.01174	.70162	2.6590	.00756
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.81483	.00610	.00617	F 411.44	.00858	.68987	.01663
Stddev	.08854	.00159	.00023	1.75	.00104	.00877	.00140
%RSD	10.866	26.091	3.7232	.42582	12.171	1.2720	8.4229
#1	.87256	.00435	.00592	410.10	.00774	.68564	.01502
#2	.71289	.00651	.00637	410.79	.00975	.69996	.01728
#3	.85904	.00745	.00621	413.42	.00825	.68401	.01758
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Sample Name: L1606099102 Acquired: 6/23/2016 19:00:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00429	-.00276	2.6532	.01212	.18283	.14118	-.00012
Stddev	.00146	.00633	.0058	.00057	.00067	.01057	.00163
%RSD	33.919	229.47	.21752	4.7117	.36652	7.4900	1330.7
#1	.00526	-.00115	2.6467	.01254	.18210	.13107	.00136
#2	.00500	.00261	2.6548	.01147	.18294	.14030	-.00187
#3	.00262	-.00973	2.6579	.01236	.18343	.15217	.00014
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.02620	.14570	6.4213
Stddev	.00048	.00024	.3214
%RSD	1.8484	.16660	5.0053
#1	.02567	.14542	6.6480
#2	.02632	.14589	6.0535
#3	.02662	.14577	6.5624
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6798.1	41761.	2199.9
Stddev	24.0	167.	13.3
%RSD	.35371	.40060	.60321
#1	6803.8	41578.	2190.0
#2	6771.7	41906.	2194.6
#3	6818.8	41798.	2214.9

Sample Name: L1606099501 Acquired: 6/23/2016 19:04:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00066	.03319	.00198	4.2612	22.007	.00013	F 2231.4
Stddev	.00073	.00812	.00497	.0108	.140	.00007	10.9
%RSD	111.66	24.468	250.83	.25328	.63583	52.114	.49013
#1	.00010	.02381	.00383	4.2649	22.169	.00005	2238.3
#2	.00039	.03790	-.00365	4.2491	21.924	.00018	2218.8
#3	.00149	.03786	.00577	4.2697	21.929	.00017	2237.2
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail
High Limit							270.00
Low Limit							-.10000

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00230	.00542	.00437	.00761	16.049	77.831	15.290
Stddev	.00066	.00043	.00087	.00268	.109	.457	.102
%RSD	28.682	8.0139	19.948	35.207	.68074	.58771	.66563
#1	.00204	.00521	.00375	.01070	16.175	78.291	15.407
#2	.00305	.00592	.00537	.00603	15.989	77.376	15.216
#3	.00181	.00514	.00400	.00609	15.983	77.827	15.248
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	289.46	2.5248	-.00161	F 5161.1	-.00605	-.04671	-.00084
Stddev	1.34	.0130	.00040	91.9	.00062	.01060	.00404
%RSD	.46449	.51604	24.821	1.7808	10.259	22.695	480.02
#1	290.80	2.5348	-.00118	5246.4	-.00638	-.03903	-.00094
#2	288.11	2.5101	-.00167	5063.8	-.00533	-.05880	.00324
#3	289.46	2.5296	-.00197	5173.3	-.00644	-.04229	-.00483
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Sample Name: L1606099501 Acquired: 6/23/2016 19:04:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00001**	**.00060**	**1.7140**	**-.00270**	F **293.43**	F **-.23021**	**-.00582**
Stddev	.00390	.00430	.0091	.00145	3.55	.00343	.00281
%RSD	37587.	722.43	.53043	53.583	1.2096	1.4919	48.318
#1	.00450	-.00436	1.7206	-.00436	292.37	-.23049	-.00302
#2	-.00251	.00275	1.7036	-.00173	297.39	-.22665	-.00864
#3	-.00196	.00340	1.7177	-.00201	290.53	-.23350	-.00579
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Fail	Chk Pass
High Limit					9.0000	36.000	
Low Limit					-.01000	-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00391**	**.46449**	F **-.84964**
Stddev	.00112	.00152	.26917
%RSD	28.498	.32684	31.680
#1	-.00485	.46570	-1.1604
#2	-.00421	.46278	-.69319
#3	-.00268	.46498	-.69528
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**5043.9**	**30664.**	**1834.7**
Stddev	16.9	150.	7.6
%RSD	.33492	.48802	.41407
#1	5024.6	30608.	1826.5
#2	5056.0	30834.	1836.3
#3	5051.1	30550.	1841.4

Sample Name: L1606099502 Acquired: 6/23/2016 19:08:29 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00345**	**.02306**	**-.00311**	**4.2770**	**21.493**	**.00016**	F **2199.4**
Stddev	.00099	.01662	.00473	.0112	.107	.00010	10.2
%RSD	28.739	72.096	152.37	.26279	.49855	65.563	.46331
#1	-.00260	.03568	.00094	4.2665	21.584	.00005	2207.7
#2	-.00454	.00422	-.00194	4.2758	21.375	.00025	2202.6
#3	-.00321	.02928	-.00831	4.2888	21.521	.00017	2188.0
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail
High Limit							270.00
Low Limit							-.10000

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00258**	**.00505**	**.00245**	**.01034**	**14.967**	**76.431**	**15.054**
Stddev	.00018	.00018	.00130	.00137	.073	.482	.075
%RSD	6.8313	3.5760	52.982	13.296	.48454	.63040	.50059
#1	.00238	.00511	.00191	.00878	14.985	76.519	15.096
#2	.00271	.00519	.00151	.01138	14.888	75.911	14.967
#3	.00265	.00485	.00393	.01085	15.029	76.862	15.100
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**285.71**	**2.4897**	**-.00104**	F **4892.3**	**-.00500**	**-.05393**	**.00746**
Stddev	1.59	.0149	.00055	124.5	.00081	.00583	.00104
%RSD	.55735	.59915	52.735	2.5451	16.181	10.801	13.884
#1	287.47	2.5069	-.00064	4942.1	-.00421	-.06066	.00665
#2	284.37	2.4802	-.00081	4984.2	-.00495	-.05038	.00710
#3	285.29	2.4820	-.00167	4750.6	-.00582	-.05076	.00863
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Sample Name: L1606099502 Acquired: 6/23/2016 19:08:29 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00124**	**.00073**	**1.5901**	**-.00205**	F **282.05**	F **-.22951**	**-.00301**
Stddev	.00214	.00258	.0105	.00077	1.08	.01217	.00231
%RSD	172.21	351.60	.66137	37.499	.38242	5.3030	76.703
#1	.00091	.00117	1.5782	-.00289	282.22	-.24190	-.00355
#2	-.00128	.00306	1.5942	-.00138	283.02	-.21757	-.00499
#3	-.00337	-.00204	1.5980	-.00189	280.89	-.22904	-.00048
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Fail	Chk Pass
High Limit					9.0000	36.000	
Low Limit					-.01000	-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00411**	**.45318**	F **-.64740**
Stddev	.00129	.00131	.07796
%RSD	31.297	.28971	12.042
#1	-.00333	.45167	-.58664
#2	-.00340	.45405	-.73530
#3	-.00559	.45381	-.62026
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**5048.8**	**30332.**	**1858.4**
Stddev	3.9	253.	17.7
%RSD	.07715	.83404	.95240
#1	5053.2	30378.	1844.8
#2	5046.9	30059.	1852.1
#3	5046.2	30558.	1878.4

Sample Name: L1606106401 Acquired: 6/23/2016 19:12:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00137	.93109	.00053	.03729	.05113	.00000	28.407	.00020
Stddev	.00252	.00902	.00338	.00153	.00202	.00011	.262	.00043
%RSD	183.56	.96930	639.15	4.0949	3.9481	2244.2	.92307	214.59
#1	-.00121	.93101	.00099	.03594	.05319	.00011	28.651	.00015
#2	.00151	.94015	.00365	.03894	.05104	.00001	28.439	-.00020
#3	.00382	.92210	-.00306	.03698	.04915	-.00011	28.129	.00066
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00058	.00021	.00488	1.1389	1.0564	-.00108	2.9741	.01385
Stddev	.00032	.00086	.00142	.0345	.0930	.00288	.1109	.00248
%RSD	56.184	402.03	29.165	3.0303	8.7985	268.17	3.7294	17.924
#1	.00057	.00116	.00392	1.1770	1.0142	.00149	3.1007	.01663
#2	.00025	-.00002	.00421	1.1096	.99211	-.00052	2.9273	.01185
#3	.00090	-.00050	.00652	1.1303	1.1630	-.00420	2.8942	.01308
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00103	3.1622	.00158	.08777	.00276	.00045	-.00572	4.3193
Stddev	.00012	.1302	.00086	.00354	.00085	.00135	.01212	.0046
%RSD	11.727	4.1165	54.479	4.0300	30.606	297.83	212.07	.10550
#1	.00117	3.2928	.00230	.08694	.00254	-.00043	.00769	4.3152
#2	.00097	3.1615	.00063	.08472	.00205	.00201	-.01591	4.3242
#3	.00095	3.0324	.00181	.09165	.00370	-.00022	-.00892	4.3184
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606106401 Acquired: 6/23/2016 19:12:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00060	.09591	.01130	-.00267	.00099	.00793	.56831
Stddev	.00047	.00237	.00917	.00175	.00090	.00003	.52710
%RSD	78.603	2.4713	81.158	65.518	91.033	.34720	92.749
#1	.00074	.09837	.00228	-.00301	.00008	.00796	.23401
#2	.00098	.09572	.01099	-.00422	.00188	.00791	.29499
#3	.00007	.09364	.02061	-.00077	.00100	.00793	1.1759
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6738.4	41869.	2132.0
Stddev	11.1	208.	5.4
%RSD	.16498	.49610	.25526
#1	6728.6	41753.	2129.0
#2	6736.2	41746.	2138.3
#3	6750.5	42109.	2128.7

Sample Name: CCV Acquired: 6/23/2016 19:16:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.37919	9.4539	.37838	.47741	.95410	.04762	9.5636
Stddev	.00221	.0169	.00039	.00245	.00513	.00013	.0372
%RSD	.58374	.17847	.10355	.51410	.53813	.27668	.38859
#1	.38055	9.4678	.37878	.47468	.94817	.04774	9.5214
#2	.38039	9.4588	.37835	.47943	.95695	.04748	9.5914
#3	.37664	9.4351	.37800	.47812	.95718	.04762	9.5780
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04743	.19073	.47732	.47745	3.7981	47.840	.97103
Stddev	.00050	.00017	.00127	.00151	.0352	.199	.00699
%RSD	1.0527	.08963	.26621	.31598	.92728	.41493	.71955
#1	.04779	.19083	.47845	.47601	3.7576	47.614	.96329
#2	.04763	.19053	.47595	.47902	3.8152	47.921	.97687
#3	.04686	.19083	.47757	.47730	3.8215	47.986	.97295
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.5182	.47503	.95370	48.127	.47728	9.3799	.47857
Stddev	.1107	.00550	.00396	.203	.00099	.0039	.00307
%RSD	1.1627	1.1583	.41501	.42134	.20780	.04109	.64155
#1	9.6428	.46909	.95745	47.914	.47781	9.3770	.47874
#2	9.4313	.47605	.95408	48.150	.47790	9.3843	.48155
#3	9.4805	.47995	.94956	48.317	.47614	9.3784	.47541
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 19:16:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1355**	**.37785**	**4.7735**	**.94584**	**.95950**	**.94974**	**.48314**
Stddev	.0014	.00604	.0057	.00205	.00431	.00481	.00202
%RSD	.12258	1.5984	.11925	.21676	.44891	.50688	.41802
#1	1.1369	.38109	4.7670	.94374	.95466	.94648	.48517
#2	1.1341	.38158	4.7762	.94595	.96094	.95527	.48113
#3	1.1355	.37088	4.7773	.94784	.96291	.94747	.48313
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.95446**	**.95114**	F **.47272**
Stddev	.00286	.00014	.27938
%RSD	.29988	.01520	59.100
#1	.95768	.95118	.66932
#2	.95222	.95127	.15292
#3	.95348	.95099	.59591
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6901.5**	**42860.**	**2195.1**
Stddev	41.4	316.	16.1
%RSD	.60057	.73673	.73557
#1	6924.3	42538.	2183.7
#2	6853.6	42875.	2188.1
#3	6926.5	43169.	2213.6

Sample Name: CCB Acquired: 6/23/2016 19:20:08 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00238**	**-.01466**	**.00143**	**.00483**	**-.00003**	**.00001**	**.07786**
Stddev	.00049	.00200	.00149	.00453	.00214	.00009	.06354
%RSD	20.427	13.611	104.06	93.858	6527.4	940.84	81.613
#1	.00284	-.01261	.00199	.00972	.00118	.00006	.15097
#2	.00187	-.01659	-.00026	.00400	.00122	.00006	.03598
#3	.00244	-.01478	.00255	.00077	-.00250	-.00009	.04662
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00019**	**-.00031**	**.00022**	**.00039**	**.00442**	**.00595**	**.00280**
Stddev	.00007	.00052	.00124	.00132	.02693	.06585	.00064
%RSD	37.292	170.19	563.67	338.25	609.56	1107.4	22.926
#1	.00023	-.00057	.00004	-.00106	.01159	.04036	.00265
#2	.00011	-.00065	-.00092	.00069	-.02537	-.06998	.00350
#3	.00023	.00030	.00154	.00154	.02703	.04745	.00224
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.03600**	**.00167**	**.00345**	**.42320**	**-.00024**	**.00139**	F **-.00578**
Stddev	.09635	.00312	.00061	.03957	.00049	.00443	.00184
%RSD	267.61	187.01	17.568	9.3512	202.30	317.35	31.895
#1	.01890	-.00154	.00285	.45813	-.00079	.00606	-.00650
#2	.13975	.00185	.00406	.43125	.00015	.00087	-.00715
#3	-.05065	.00469	.00345	.38022	-.00009	-.00274	-.00368
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail
High Limit							.00500
Low Limit							-.00500

Sample Name: CCB Acquired: 6/23/2016 19:20:08 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00539**	**-.00364**	**.00090**	**-.00057**	**.00373**	**-.00314**	**.00032**
Stddev	.00515	.00131	.00038	.00067	.00020	.00469	.00190
%RSD	95.593	36.045	42.719	116.62	5.2758	149.60	585.89
#1	.01074	-.00478	.00116	-.00075	.00392	.00167	.00249
#2	.00495	-.00221	.00108	.00016	.00352	-.00771	-.00042
#3	.00047	-.00394	.00046	-.00113	.00373	-.00337	-.00110
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00043**	**.00005**	F **.04542**
Stddev	.00072	.00012	.34069
%RSD	169.53	237.68	750.04
#1	-.00008	-.00001	.43855
#2	-.00125	.00019	-.16363
#3	.00006	-.00003	-.13865
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6967.5**	**43395.**	**2190.0**
Stddev	41.4	704.	41.3
%RSD	.59409	1.6225	1.8846
#1	6953.3	43361.	2223.7
#2	6935.1	44115.	2144.0
#3	7014.1	42708.	2202.3

Sample Name: PBW MC Acquired: 6/23/2016 19:24:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00023	.01182	-.00092	.00493	-.00012	-.00010	.03899	.00002
Stddev	.00163	.00860	.00408	.00276	.00157	.00005	.01101	.00042
%RSD	702.25	72.769	445.47	56.096	1293.5	54.721	28.229	1843.4
#1	.00065	.00253	-.00537	.00388	.00151	-.00014	.04722	.00048
#2	.00161	.01951	.00264	.00806	-.00023	-.00011	.02649	-.00036
#3	-.00157	.01343	-.00001	.00284	-.00164	-.00004	.04326	-.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00017	.00049	.00204	.00432	.17133	-.00238	.06048	.00264
Stddev	.00029	.00050	.00040	.03628	.08746	.00474	.03409	.00131
%RSD	168.82	101.29	19.724	840.04	51.052	199.46	56.354	49.484
#1	-.00010	.00067	.00239	-.02516	.27127	.00176	.03814	.00248
#2	-.00049	.00088	.00214	-.00672	.10876	-.00754	.09972	.00142
#3	.00007	-.00007	.00160	.04484	.13394	-.00135	.04360	.00402
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00096	.32980	.00109	-.00655	-.00048	.00411	-.00330	-.02137
Stddev	.00036	.05000	.00076	.00347	.00344	.00130	.00992	.00312
%RSD	37.319	15.161	69.359	52.886	720.16	31.721	300.60	14.588
#1	.00079	.36140	.00194	-.01016	-.00415	.00417	-.01302	-.02373
#2	.00072	.35584	.00049	-.00325	.00266	.00538	.00681	-.02254
#3	.00137	.27215	.00084	-.00625	.00005	.00278	-.00369	-.01784
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: PBW MC Acquired: 6/23/2016 19:24:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00040**	**.00085**	**-.00321**	**-.00159**	**.00013**	**.00182**	**-.03501**
Stddev	.00061	.00058	.00748	.00275	.00086	.00009	.26030
%RSD	153.26	68.513	233.36	172.78	656.72	4.7398	743.52
#1	.00056	.00034	.00354	.00057	-.00054	.00177	-.32167
#2	.00090	.00149	-.01126	-.00468	-.00016	.00177	.03005
#3	-.00028	.00073	-.00190	-.00066	.00110	.00192	.18659
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6713.5**	**41826.**	**2105.1**
Stddev	54.0	279.	25.0
%RSD	.80485	.66739	1.1867
#1	6727.2	41946.	2094.7
#2	6654.0	42025.	2087.1
#3	6759.4	41507.	2133.7

Sample Name: LCSW MC Acquired: 6/23/2016 19:28:07 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20350	4.9935	.19725	.98212	.51929	.02489	5.3105	.02494
Stddev	.00278	.0008	.00175	.00061	.00709	.00006	.0323	.00007
%RSD	1.3668	.01694	.88835	.06228	1.3650	.26082	.60923	.28776
#1	.20050	4.9929	.19553	.98183	.52566	.02483	5.3424	.02500
#2	.20403	4.9931	.19903	.98282	.52055	.02488	5.3113	.02486
#3	.20598	4.9945	.19719	.98170	.51165	.02496	5.2777	.02497
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10152	.25256	.25516	2.0797	25.740	.52639	5.1902	.25850
Stddev	.00055	.00072	.00042	.0675	.264	.00403	.0928	.00472
%RSD	.53818	.28671	.16566	3.2456	1.0251	.76543	1.7881	1.8242
#1	.10140	.25339	.25549	2.1412	25.923	.53020	5.2421	.26354
#2	.10211	.25215	.25468	2.0904	25.858	.52680	5.2454	.25420
#3	.10104	.25212	.25531	2.0075	25.437	.52218	5.0831	.25775
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.51567	25.901	.25435	4.8264	.25513	.60348	.19064	2.5576
Stddev	.00163	.329	.00155	.0089	.00174	.00417	.00132	.0042
%RSD	.31684	1.2719	.61000	.18425	.68292	.69074	.69080	.16255
#1	.51692	26.132	.25556	4.8337	.25335	.60561	.19063	2.5545
#2	.51627	26.047	.25489	4.8291	.25683	.60615	.18933	2.5560
#3	.51382	25.523	.25260	4.8165	.25520	.59868	.19196	2.5623
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: LCSW MC Acquired: 6/23/2016 19:28:07 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50423	.52568	.51791	.25225	.50435	.50645	.50871
Stddev	.00241	.00482	.00534	.00263	.00086	.00034	.12302
%RSD	.47730	.91755	1.0320	1.0435	.16959	.06729	24.183
#1	.50649	.52965	.52307	.24922	.50405	.50684	.42537
#2	.50170	.52708	.51825	.25391	.50532	.50620	.65001
#3	.50450	.52032	.51240	.25363	.50369	.50631	.45076
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6675.8	41467.	2105.7
Stddev	24.4	77.	45.0
%RSD	.36501	.18473	2.1351
#1	6674.7	41472.	2065.5
#2	6652.0	41541.	2097.5
#3	6700.7	41388.	2154.3

Sample Name: F BLANK Acquired: 6/23/2016 19:31:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573025-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00035**	**.01697**	**.00017**	**.00823**	**.00030**	**-.00002**	**.01424**	**.00010**
Stddev	.00143	.00343	.00368	.00159	.00073	.00006	.02636	.00016
%RSD	406.65	20.219	2149.5	19.295	245.04	366.35	185.13	161.49
#1	-.00146	.01939	.00373	.00699	-.00054	.00002	.03225	-.00007
#2	-.00084	.01848	.00040	.00767	.00069	.00002	.02649	.00012
#3	.00126	.01304	-.00362	.01002	.00075	-.00008	-.01602	.00025
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00015**	**.00071**	**.00022**	**-.00250**	**.21680**	**-.00037**	**.04710**	**.00203**
Stddev	.00058	.00084	.00113	.01048	.06228	.00321	.02108	.00203
%RSD	395.96	117.23	505.81	419.76	28.728	858.57	44.755	100.05
#1	-.00038	.00061	.00120	.00643	.26535	.00293	.05892	.00057
#2	.00005	-.00007	.00049	.00011	.14658	-.00348	.02276	.00117
#3	.00076	.00160	-.00102	-.01403	.23847	-.00057	.05962	.00435
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00063**	**146.49**	**.00027**	**.00148**	**-.00007**	**-.00070**	**-.00157**	**-.01682**
Stddev	.00016	.16	.00052	.00584	.00253	.00452	.00481	.00103
%RSD	25.353	.10971	193.88	395.20	3668.9	645.60	306.24	6.1510
#1	.00048	146.42	-.00030	-.00517	.00173	.00230	.00206	-.01801
#2	.00061	146.38	.00070	.00384	-.00296	.00151	-.00702	-.01610
#3	.00079	146.68	.00040	.00577	.00102	-.00590	.00025	-.01635
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: F BLANK Acquired: 6/23/2016 19:31:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573025-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00050**	**.00113**	**.00420**	**-.00413**	**.00010**	**.00416**	**.06618**
Stddev	.00105	.00030	.00358	.00325	.00062	.00033	.30902
%RSD	208.81	26.175	85.339	78.560	609.38	7.9682	466.95
#1	.00054	.00092	.00586	-.00501	.00082	.00444	.36201
#2	.00153	.00147	.00009	-.00685	-.00027	.00424	.09104
#3	-.00056	.00099	.00664	-.00054	-.00025	.00379	-.25452
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6739.3**	**41425.**	**2117.9**
Stddev	34.7	212.	17.6
%RSD	.51557	.51131	.83142
#1	6776.6	41379.	2104.1
#2	6733.4	41656.	2137.7
#3	6707.8	41240.	2111.8

Sample Name: L1606077902 Acquired: 6/23/2016 19:35:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00041	.10466	.00285	.06045	.05584	-.00002	6.2174	.00046
Stddev	.00116	.00710	.00203	.00084	.00161	.00004	.0446	.00027
%RSD	281.92	6.7835	71.407	1.3870	2.8912	266.76	.71767	58.204
#1	-.00169	.09707	.00256	.05995	.05497	-.00004	6.2605	.00076
#2	-.00009	.10577	.00097	.05998	.05770	-.00005	6.2202	.00025
#3	.00055	.11114	.00501	.06142	.05485	.00003	6.1714	.00037
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00045	.00071	.00269	.06613	.83231	.00096	1.9162	.03669
Stddev	.00019	.00047	.00085	.02858	.05178	.00081	.1367	.00368
%RSD	41.445	66.440	31.592	43.213	6.2217	84.544	7.1320	10.038
#1	.00040	.00124	.00202	.09508	.77901	.00190	2.0666	.03505
#2	.00066	.00054	.00241	.03794	.88243	.00048	1.8822	.04090
#3	.00030	.00035	.00365	.06539	.83549	.00050	1.7997	.03410
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00023	146.89	.00039	.00511	.00718	-.00047	-.00261	.31279
Stddev	.00029	.83	.00077	.00679	.00094	.00278	.00523	.00191
%RSD	125.89	.56640	197.90	132.70	13.092	597.43	200.57	.60978
#1	.00003	147.12	.00125	.00567	.00825	-.00357	.00205	.31442
#2	.00057	147.59	-.00024	-.00193	.00647	.00180	-.00827	.31325
#3	.00010	145.97	.00016	.01160	.00682	.00038	-.00160	.31069
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606077902 Acquired: 6/23/2016 19:35:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00094**	**.03050**	**.00108**	**-.00165**	**-.00052**	**.01606**	**.48222**
Stddev	.00106	.00016	.00532	.00346	.00115	.00010	.08551
%RSD	113.09	.52207	493.59	209.68	220.39	.59709	17.733
#1	.00063	.03034	.00420	.00102	-.00167	.01612	.45722
#2	.00006	.03050	-.00506	-.00556	.00063	.01595	.57745
#3	.00212	.03066	.00410	-.00041	-.00053	.01610	.41199
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6708.2**	**41665.**	**2172.2**
Stddev	31.1	139.	27.0
%RSD	.46322	.33365	1.2422
#1	6681.0	41585.	2158.5
#2	6701.6	41584.	2154.7
#3	6742.1	41825.	2203.2

Sample Name: L1606077905S Acquired: 6/23/2016 19:39:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20228	5.2112	.20044	1.0337	.55827	.02485	9.6912	.02500
Stddev	.00166	.0132	.00213	.0032	.00578	.00010	.1284	.00041
%RSD	.82305	.25409	1.0603	.31079	1.0349	.40858	1.3248	1.6502
#1	.20387	5.2188	.19844	1.0331	.56463	.02478	9.8330	.02530
#2	.20055	5.1959	.20020	1.0371	.55682	.02480	9.6578	.02517
#3	.20241	5.2189	.20267	1.0308	.55335	.02497	9.5828	.02453
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09974	.24950	.25228	2.2679	26.280	.52442	6.0599	.27187
Stddev	.00013	.00041	.00146	.0210	.482	.00917	.1252	.00132
%RSD	.12865	.16628	.57688	.92528	1.8346	1.7485	2.0662	.48482
#1	.09988	.24904	.25061	2.2895	26.785	.53322	5.9629	.27277
#2	.09972	.24962	.25292	2.2477	26.232	.52510	6.2012	.27249
#3	.09962	.24984	.25330	2.2665	25.824	.51492	6.0156	.27036
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.51461	175.16	.24930	4.8996	.25282	.59906	.19939	3.1901
Stddev	.00031	1.57	.00247	.0075	.00056	.00386	.00642	.0017
%RSD	.06092	.89896	.99271	.15227	.22003	.64505	3.2200	.05219
#1	.51467	176.79	.25185	4.8930	.25302	.59517	.19207	3.1907
#2	.51427	175.03	.24915	4.9077	.25325	.60289	.20408	3.1882
#3	.51489	173.65	.24691	4.8982	.25219	.59912	.20203	3.1913
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606077905S Acquired: 6/23/2016 19:39:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50008	.54325	.51187	.24743	.50223	.51373	.44679
Stddev	.00071	.00377	.00743	.00180	.00201	.00029	.30975
%RSD	.14141	.69391	1.4524	.72816	.40098	.05697	69.327
#1	.49932	.54729	.51807	.24579	.50124	.51369	.24650
#2	.50021	.54262	.51392	.24714	.50091	.51403	.29031
#3	.50072	.53983	.50363	.24936	.50455	.51345	.80356
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6683.1	41005.	2151.4
Stddev	6.8	305.	41.0
%RSD	.10241	.74481	1.9080
#1	6678.0	40809.	2104.3
#2	6690.9	41357.	2179.3
#3	6680.4	40849.	2170.6

Sample Name: L1606077908SD Acquired: 6/23/2016 19:43:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-06

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20288	5.1811	.19942	1.0150	.55237	.02488	9.4634	.02476
Stddev	.00176	.0005	.00235	.0035	.00196	.00014	.0234	.00044
%RSD	.86650	.00922	1.1808	.34043	.35412	.56384	.24707	1.7705
#1	.20137	5.1816	.19728	1.0163	.55459	.02505	9.4654	.02526
#2	.20245	5.1810	.20195	1.0110	.55090	.02479	9.4391	.02451
#3	.20481	5.1807	.19904	1.0175	.55161	.02482	9.4857	.02450
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10026	.24869	.25215	2.1544	26.013	.51658	6.1704	.26979
Stddev	.00032	.00041	.00151	.0430	.062	.00117	.0894	.00386
%RSD	.31797	.16528	.59908	1.9976	.23856	.22630	1.4482	1.4295
#1	.10042	.24885	.25049	2.2010	26.026	.51657	6.2344	.26873
#2	.10048	.24900	.25345	2.1162	25.946	.51776	6.2085	.27407
#3	.09990	.24822	.25251	2.1458	26.068	.51542	6.0683	.26657
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.51646	167.22	.25037	4.9060	.24566	.60333	.19310	3.1314
Stddev	.00260	.31	.00182	.0167	.00482	.00537	.00965	.0091
%RSD	.50438	.18479	.72893	.34087	1.9619	.89072	4.9996	.29092
#1	.51697	167.57	.24950	4.9167	.24626	.60535	.20424	3.1326
#2	.51877	167.00	.25246	4.9146	.25015	.60741	.18713	3.1398
#3	.51363	167.09	.24914	4.8868	.24056	.59724	.18793	3.1217
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606077908SD Acquired: 6/23/2016 19:43:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-06

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49955**	**.54302**	**.51966**	**.24675**	**.50221**	**.51185**	**.71645**
Stddev	.00029	.00244	.00212	.00186	.00227	.00193	.20134
%RSD	.05847	.45013	.40726	.75321	.45223	.37767	28.103
#1	.49934	.54570	.52008	.24714	.50300	.51112	.72584
#2	.49989	.54245	.52153	.24839	.49965	.51404	.51058
#3	.49944	.54091	.51736	.24473	.50398	.51039	.91294
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6680.0**	**41465.**	**2146.4**
Stddev	27.3	183.	22.0
%RSD	.40866	.44237	1.0253
#1	6690.0	41465.	2121.9
#2	6649.1	41648.	2153.0
#3	6700.8	41281.	2164.4

Sample Name: L1606077911 Acquired: 6/23/2016 19:47:01 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00061**	**.17290**	**.00164**	**.05825**	**.04624**	**.00003**	**4.2798**
Stddev	.00118	.00586	.00116	.00187	.00146	.00004	.0673
%RSD	194.32	3.3883	71.027	3.2055	3.1581	127.98	1.5735
#1	-.00071	.17964	.00069	.05609	.04622	.00002	4.3575
#2	.00096	.16999	.00128	.05931	.04771	.00007	4.2378
#3	.00156	.16906	.00294	.05934	.04479	-.00001	4.2442
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00024**	**.00029**	**-.00045**	**.00174**	**.09875**	**.44424**	**.00053**
Stddev	.00014	.00027	.00192	.00071	.01625	.12676	.00083
%RSD	58.518	92.659	422.78	40.567	16.451	28.534	157.87
#1	.00014	.00059	.00168	.00219	.10612	.58035	.00098
#2	.00017	.00020	-.00202	.00093	.11000	.42278	-.00043
#3	.00040	.00008	-.00103	.00212	.08013	.32958	.00104
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.72600**	**.00518**	**.00043**	**142.06**	**.00228**	**.00480**	**-.00087**
Stddev	.03723	.00162	.00026	.88	.00064	.00070	.00109
%RSD	5.1278	31.190	61.921	.61608	28.046	14.500	124.44
#1	.71844	.00688	.00028	143.07	.00166	.00434	.00015
#2	.76644	.00366	.00073	141.60	.00293	.00560	-.00076
#3	.69314	.00502	.00026	141.51	.00224	.00446	-.00201
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606077911 Acquired: 6/23/2016 19:47:01 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00404**	**-.00935**	**.37687**	**.00027**	**.01793**	**.00337**	**-.00570**
Stddev	.00215	.00588	.00275	.00092	.00024	.00451	.00216
%RSD	53.318	62.841	.73023	343.85	1.3281	134.02	37.974
#1	.00159	-.01397	.37384	-.00077	.01800	-.00085	-.00756
#2	.00487	-.00274	.37757	.00097	.01812	.00283	-.00332
#3	.00566	-.01135	.37921	.00060	.01766	.00813	-.00622
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00079**	**.00664**	F **-.09970**
Stddev	.00084	.00026	.16681
%RSD	107.32	3.9795	167.31
#1	-.00163	.00690	-.09910
#2	.00005	.00665	.06681
#3	-.00078	.00638	-.26680
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6697.8**	**41398.**	**2131.1**
Stddev	24.2	144.	21.0
%RSD	.36115	.34894	.98383
#1	6693.3	41435.	2107.1
#2	6723.9	41521.	2140.2
#3	6676.1	41239.	2145.9

Sample Name: L1606077911PS Acquired: 6/23/2016 19:50:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573973-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20166	5.1361	.19947	1.0226	.54909	.02495	8.9751	.02473
Stddev	.00114	.0195	.00083	.0025	.00185	.00016	.0498	.00035
%RSD	.56716	.37888	.41683	.24775	.33639	.63847	.55502	1.4285
#1	.20230	5.1582	.20042	1.0247	.55108	.02513	8.9915	.02440
#2	.20234	5.1215	.19889	1.0198	.54877	.02481	9.0146	.02510
#3	.20034	5.1285	.19910	1.0233	.54742	.02492	8.9192	.02469
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10039	.24897	.25158	2.1225	25.565	.51154	5.6600	.25879
Stddev	.00018	.00163	.00180	.0342	.089	.00200	.0479	.00217
%RSD	.17967	.65380	.71732	1.6127	.34620	.39088	.84555	.83680
#1	.10051	.24824	.25076	2.1608	25.630	.51077	5.6851	.25647
#2	.10018	.24783	.25365	2.0948	25.465	.51005	5.6902	.26076
#3	.10047	.25083	.25032	2.1119	25.602	.51382	5.6048	.25913
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.51470	153.34	.25033	4.8965	.24793	.59559	.19243	2.8850
Stddev	.00123	.64	.00055	.0032	.00147	.00310	.01290	.0015
%RSD	.23864	.41616	.22124	.06430	.59145	.52110	6.7059	.05201
#1	.51611	153.87	.25059	4.8983	.24682	.59605	.19308	2.8838
#2	.51410	153.52	.25070	4.8929	.24959	.59844	.17922	2.8846
#3	.51388	152.63	.24969	4.8984	.24737	.59228	.20500	2.8867
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606077911PS Acquired: 6/23/2016 19:50:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573973-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50192	.53133	.51057	.24460	.50225	.50838	.44531
Stddev	.00111	.00315	.01429	.00114	.00111	.00054	.18642
%RSD	.22110	.59328	2.7983	.46703	.22030	.10637	41.862
#1	.50147	.53348	.52543	.24579	.50343	.50865	.55546
#2	.50112	.53279	.49693	.24352	.50209	.50873	.55041
#3	.50319	.52771	.50934	.24448	.50123	.50775	.23008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6651.9	41058.	2146.9
Stddev	34.9	202.	7.0
%RSD	.52428	.49305	.32439
#1	6618.2	41118.	2148.4
#2	6649.8	40833.	2139.3
#3	6687.9	41224.	2153.0

Sample Name: L1606077911SDL Acquired: 6/23/2016 19:54:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573973-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00006**	**.03191**	**.00319**	**.01017**	**.00812**	**.00002**	**.79168**
Stddev	.00337	.01356	.00266	.00277	.00094	.00006	.04789
%RSD	5775.7	42.477	83.503	27.198	11.538	281.43	6.0487
#1	.00377	.04752	.00625	.01189	.00906	.00005	.74079
#2	-.00077	.02310	.00144	.01163	.00718	-.00005	.83586
#3	-.00282	.02511	.00188	.00698	.00812	.00006	.79839
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00007**	**.00020**	**-.00102**	**.00161**	**.02774**	**.12243**	**.00086**
Stddev	.00008	.00022	.00131	.00175	.00237	.07914	.00524
%RSD	112.15	111.56	128.35	108.58	8.5304	64.641	611.10
#1	.00015	-.00005	-.00253	-.00029	.02829	.21145	-.00473
#2	-.00001	.00027	-.00015	.00315	.02514	.09584	.00164
#3	.00009	.00038	-.00039	.00199	.02978	.06001	.00566
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05770**	**.00304**	**.00047**	**25.965**	**-.00008**	**.01273**	**.00119**
Stddev	.09190	.00232	.00015	.345	.00131	.01585	.00647
%RSD	159.28	76.323	31.599	1.3295	1579.4	124.52	542.72
#1	.15485	.00380	.00035	25.792	-.00140	.01160	.00004
#2	.04609	.00044	.00064	25.741	-.00007	.02911	-.00462
#3	-.02785	.00489	.00043	26.363	.00122	-.00253	.00817
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606077911SDL Acquired: 6/23/2016 19:54:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573973-02

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00030**	**-.00465**	**.06796**	**.00020**	**.00397**	**-.00466**	**-.00173**
Stddev	.00330	.00198	.00111	.00072	.00038	.00590	.00388
%RSD	1090.6	42.606	1.6393	364.61	9.6242	126.80	224.33
#1	.00173	-.00263	.06819	.00039	.00374	-.01146	-.00420
#2	.00148	-.00660	.06894	-.00060	.00441	-.00094	-.00374
#3	-.00411	-.00472	.06675	.00080	.00377	-.00157	.00274
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.00000**	**.00179**	F **-.14942**
Stddev	.00024	.00022	.27065
%RSD	17591.	12.069	181.14
#1	-.00010	.00175	-.05629
#2	-.00017	.00203	-.45434
#3	.00028	.00160	.06237
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6956.7**	**42908.**	**2182.4**
Stddev	24.4	138.	8.1
%RSD	.35082	.32162	.37002
#1	6945.3	43022.	2184.2
#2	6984.7	42948.	2173.6
#3	6940.1	42755.	2189.5

Sample Name: L1606077911SDL Acquired: 6/23/2016 19:58:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573973-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00091**	**.00551**	**-.00163**	**.00298**	**.00016**	**-.00004**	**.14622**	**-.00004**
Stddev	.00035	.00230	.00441	.00295	.00019	.00009	.00718	.00022
%RSD	37.880	41.654	270.73	98.999	121.74	255.43	4.9086	633.58
#1	-.00082	.00812	-.00328	.00338	.00020	.00005	.14190	-.00029
#2	-.00130	.00463	.00337	-.00015	.00033	-.00003	.14227	.00014
#3	-.00062	.00379	-.00498	.00571	-.00005	-.00013	.15451	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**-.00012**	**.00129**	**.01297**	**.07882**	**-.00139**	**.02445**	**.00297**
Stddev	.00029	.00066	.00139	.04085	.05847	.00356	.09577	.00349
%RSD	220.77	550.63	108.07	315.10	74.183	257.22	391.63	117.45
#1	-.00041	-.00064	.00208	-.02649	.01827	-.00070	-.02314	-.00101
#2	.00018	-.00034	.00210	.01030	.08323	.00179	-.03820	.00442
#3	-.00016	.00062	-.00032	.05509	.13495	-.00524	.13470	.00550
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00006**	**4.6601**	**.00016**	**.01099**	**-.00374**	**-.00006**	**-.00666**	**-.00121**
Stddev	.00064	.0487	.00101	.00932	.00214	.00271	.00377	.00406
%RSD	1145.3	1.0448	617.86	84.814	57.083	4166.1	56.580	334.61
#1	-.00077	4.6317	.00117	.02130	-.00354	.00054	-.01063	.00307
#2	.00017	4.7163	-.00086	.00316	-.00172	-.00302	-.00622	-.00502
#3	.00043	4.6322	.00019	.00850	-.00597	.00229	-.00313	-.00169
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606077911SDL Acquired: 6/23/2016 19:58:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573973-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00037**	**.00016**	**.00324**	**-.00190**	**.00015**	**.00102**	**.45592**
Stddev	.00090	.00024	.00774	.00281	.00036	.00011	.15548
%RSD	242.24	152.87	238.78	147.46	242.78	10.924	34.103
#1	-.00010	-.00008	-.00569	-.00173	-.00002	.00104	.63162
#2	.00036	.00041	.00759	.00081	.00057	.00089	.33613
#3	-.00138	.00014	.00782	-.00479	-.00009	.00111	.40000
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**7002.1**	**43421.**	**2180.6**
Stddev	12.5	150.	18.1
%RSD	.17813	.34564	.82992
#1	7007.5	43338.	2178.3
#2	6987.9	43594.	2163.7
#3	7011.0	43331.	2199.7

Sample Name: CCV Acquired: 6/23/2016 20:02:38 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.38352	9.5676	.38149	.48360	.95868	.04827	9.5943
Stddev	.00322	.0401	.00429	.00105	.00492	.00022	.0294
%RSD	.83982	.41946	1.1243	.21691	.51367	.46090	.30624
#1	.38657	9.6123	.37703	.48417	.95340	.04832	9.5754
#2	.38382	9.5347	.38559	.48239	.95948	.04803	9.6282
#3	.38015	9.5558	.38184	.48424	.96315	.04846	9.5794
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04762	.19248	.48389	.48328	3.8584	48.266	.97868
Stddev	.00015	.00028	.00059	.00212	.0428	.089	.00720
%RSD	.31413	.14541	.12117	.43944	1.1080	.18417	.73546
#1	.04779	.19230	.48370	.48346	3.8177	48.266	.97363
#2	.04758	.19280	.48342	.48107	3.8546	48.177	.97549
#3	.04750	.19234	.48455	.48530	3.9029	48.355	.98692
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.6513	.48380	.96301	47.927	.48135	9.5102	.48331
Stddev	.2636	.00656	.00332	.072	.00114	.0199	.00231
%RSD	2.7313	1.3558	.34426	.15061	.23638	.20964	.47887
#1	9.3550	.47806	.96658	47.845	.48232	9.5216	.48282
#2	9.8597	.48238	.96243	47.981	.48165	9.4872	.48129
#3	9.7393	.49095	.96003	47.956	.48010	9.5218	.48584
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 20:02:38 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1479**	**.38358**	**4.8260**	**.95697**	**.96108**	**.96615**	**.48705**
Stddev	.0066	.01108	.0032	.00258	.00341	.00961	.00179
%RSD	.57598	2.8894	.06586	.26977	.35526	.99472	.36795
#1	1.1549	.37665	4.8269	.95791	.95718	.95798	.48502
#2	1.1417	.39636	4.8286	.95405	.96254	.96374	.48839
#3	1.1472	.37773	4.8224	.95895	.96352	.97674	.48775
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.96626**	**.96309**	F **.37739**
Stddev	.00265	.00068	.17080
%RSD	.27390	.07033	45.257
#1	.96486	.96244	.28567
#2	.96461	.96302	.27206
#3	.96931	.96379	.57445
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6922.0**	**42512.**	**2201.6**
Stddev	21.1	353.	11.3
%RSD	.30486	.83050	.51529
#1	6934.6	42346.	2213.8
#2	6933.8	42273.	2199.8
#3	6897.7	42918.	2191.3

Sample Name: CCB Acquired: 6/23/2016 20:06:16 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00019	-.00041	.00090	.00065	-.00017	-.00009	.03364	-.00010
Stddev	.00250	.00080	.00202	.00197	.00087	.00003	.02458	.00030
%RSD	1325.3	193.76	223.31	305.36	513.01	38.963	73.061	282.77
#1	-.00159	-.00083	-.00002	.00192	-.00117	-.00006	.00559	-.00032
#2	.00304	.00051	-.00049	.00164	.00038	-.00008	.04390	.00023
#3	-.00089	-.00092	.00322	-.00163	.00029	-.00012	.05142	-.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00011	.00089	.00107	.02622	.14403	.00140	.16533	.00084
Stddev	.00031	.00098	.00055	.02571	.09143	.00408	.10839	.00123
%RSD	276.41	109.41	51.427	98.056	63.481	291.75	65.560	147.25
#1	.00047	.00000	.00094	-.00236	.13572	.00290	.28604	.00006
#2	-.00006	.00194	.00060	.03356	.05704	.00451	.07632	.00226
#3	-.00008	.00074	.00168	.04747	.23934	-.00322	.13363	.00020
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00296	.11509	-.00013	.00085	.00094	.00629	-.00145	.00187
Stddev	.00021	.03731	.00042	.00219	.00279	.00604	.00094	.00124
%RSD	7.1713	32.413	339.72	257.22	298.65	96.053	64.718	66.434
#1	.00285	.13809	-.00049	.00212	.00095	.00557	-.00045	.00293
#2	.00321	.07205	.00034	.00212	.00372	.01265	-.00231	.00218
#3	.00283	.13514	-.00022	-.00168	-.00186	.00064	-.00160	.00050
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: CCB Acquired: 6/23/2016 20:06:16 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00007**	**-.00022**	**-.00145**	**-.00171**	**-.00068**	**.00008**	**.01839**
Stddev	.00140	.00022	.01047	.00376	.00047	.00019	.18176
%RSD	1871.4	101.49	723.55	219.98	69.299	243.52	988.14
#1	-.00163	-.00007	.01060	.00231	-.00016	.00013	.02617
#2	.00109	-.00047	-.00838	-.00514	-.00082	.00023	.19614
#3	.00032	-.00011	-.00656	-.00229	-.00108	-.00013	-.16713
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6970.0**	**43262.**	**2192.2**
Stddev	74.7	174.	9.5
%RSD	1.0710	.40250	.43124
#1	7003.8	43124.	2200.3
#2	7021.8	43458.	2194.3
#3	6884.5	43203.	2181.8

Sample Name: L1606077914 Acquired: 6/23/2016 20:10:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00149	.10628	.00096	.02840	.05288	-.00005	7.4368	.00027
Stddev	.00097	.00356	.00214	.00184	.00085	.00008	.1298	.00005
%RSD	64.971	3.3536	223.67	6.4703	1.6095	158.12	1.7453	17.347
#1	.00054	.10724	-.00146	.02628	.05351	-.00006	7.5825	.00032
#2	.00248	.10234	.00174	.02939	.05191	-.00012	7.3941	.00025
#3	.00145	.10927	.00259	.02952	.05322	.00003	7.3337	.00024
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00039	.00052	.00185	.08486	.70361	-.00341	1.3249	.01531
Stddev	.00024	.00198	.00037	.02791	.04945	.00257	.0384	.00314
%RSD	62.961	378.91	20.059	32.893	7.0277	75.397	2.8943	20.533
#1	.00050	.00276	.00177	.06380	.69565	-.00073	1.3540	.01542
#2	.00055	-.00104	.00226	.11652	.75656	-.00364	1.3392	.01211
#3	.00011	-.00015	.00153	.07425	.65863	-.00587	1.2814	.01840
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	136.97	.00087	.00420	.00004	.00045	-.00208	.33477
Stddev	.00023	.63	.00055	.00550	.00261	.00047	.00600	.00123
%RSD	39.437	.45874	63.261	131.00	7265.2	105.14	288.25	.36659
#1	.00032	137.68	.00147	.00987	-.00284	.00003	.00096	.33486
#2	.00071	136.77	.00073	.00385	.00070	.00095	.00179	.33594
#3	.00074	136.47	.00040	-.00112	.00225	.00036	-.00900	.33350
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606077914 Acquired: 6/23/2016 20:10:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00035	.03351	-.00005	-.00700	-.00055	.00972	.13105
Stddev	.00062	.00057	.01222	.00229	.00057	.00011	.18984
%RSD	177.30	1.6905	24453.	32.742	102.06	1.1714	144.86
#1	-.00033	.03393	.00766	-.00548	.00007	.00981	.20911
#2	.00048	.03373	-.01414	-.00589	-.00071	.00959	.26943
#3	.00090	.03286	.00633	-.00964	-.00102	.00975	-.08538
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6756.2	41698.	2168.9
Stddev	15.0	223.	10.5
%RSD	.22178	.53392	.48265
#1	6768.0	41441.	2173.7
#2	6739.3	41816.	2176.1
#3	6761.2	41837.	2156.9

Sample Name: L1606080801 Acquired: 6/23/2016 20:14:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.00981	.00166	.02700	.00536	-.00004	53.221	.00023
Stddev	.00132	.00387	.00350	.00355	.00074	.00005	.458	.00031
%RSD	223.52	39.428	210.37	13.162	13.862	133.64	.86119	137.26
#1	-.00079	.01246	.00412	.02618	.00458	.00002	53.750	-.00010
#2	.00184	.00537	-.00234	.03090	.00606	-.00008	52.944	.00025
#3	.00072	.01160	.00322	.02393	.00543	-.00006	52.968	.00052
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00007	-.00019	.00217	.26849	1.4108	-.00252	15.011	.12291
Stddev	.00031	.00023	.00068	.03341	.0586	.00219	.184	.00135
%RSD	424.72	122.66	31.151	12.444	4.1556	86.964	1.2259	1.0966
#1	-.00007	-.00003	.00235	.24539	1.3436	-.00503	15.179	.12150
#2	-.00038	-.00008	.00142	.25327	1.4510	-.00152	15.040	.12419
#3	.00023	-.00045	.00273	.30679	1.4380	-.00101	14.814	.12302
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00051	15.947	.00061	.00603	.00298	.00330	-.00384	5.1133
Stddev	.00030	.171	.00079	.00285	.00319	.00406	.01618	.0046
%RSD	57.821	1.0703	128.91	47.356	107.05	123.06	420.92	.09015
#1	.00017	16.102	-.00004	.00833	.00524	-.00067	-.02178	5.1184
#2	.00071	15.975	.00148	.00284	.00438	.00312	.00059	5.1096
#3	.00066	15.764	.00039	.00691	-.00067	.00744	.00965	5.1119
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606080801 Acquired: 6/23/2016 20:14:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00027**	**.17284**	**-.01553**	**-.00411**	**-.00044**	**.00368**	**.21621**
Stddev	.00054	.00134	.00554	.00196	.00032	.00035	.15335
%RSD	203.08	.77573	35.693	47.631	72.739	9.4159	70.926
#1	-.00035	.17423	-.01858	-.00259	-.00007	.00340	.36753
#2	.00031	.17155	-.00913	-.00632	-.00067	.00406	.22018
#3	-.00077	.17274	-.01887	-.00342	-.00059	.00356	.06091
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6663.3**	**41709.**	**2149.6**
Stddev	10.8	204.	16.5
%RSD	.16157	.48945	.76736
#1	6671.1	41908.	2131.3
#2	6667.7	41719.	2154.5
#3	6651.0	41500.	2163.2

Sample Name: L1606080803S Acquired: 6/23/2016 20:18:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20531	4.9958	.20276	1.0092	.51435	.02483	57.046	.02496
Stddev	.00292	.0043	.00269	.0024	.00665	.00006	.768	.00030
%RSD	1.4202	.08567	1.3273	.24307	1.2934	.23022	1.3467	1.1942
#1	.20859	4.9909	.20264	1.0074	.52180	.02478	57.925	.02528
#2	.20436	4.9977	.20014	1.0120	.50900	.02489	56.503	.02470
#3	.20300	4.9987	.20552	1.0083	.51225	.02482	56.709	.02490
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09967	.25096	.25022	2.3265	26.678	.52152	20.005	.37815
Stddev	.00055	.00036	.00033	.0587	.264	.00877	.198	.00476
%RSD	.55302	.14274	.13074	2.5243	.98845	1.6818	.98772	1.2593
#1	.10016	.25128	.25042	2.3809	26.980	.53116	20.035	.38218
#2	.09976	.25104	.24984	2.2643	26.497	.51402	19.794	.37290
#3	.09907	.25057	.25040	2.3343	26.556	.51939	20.186	.37938
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.52098	41.169	.25038	4.9597	.24975	.60811	.19088	7.6345
Stddev	.00121	.394	.00120	.0047	.00531	.00285	.00510	.0102
%RSD	.23295	.95680	.47973	.09548	2.1271	.46887	2.6703	.13352
#1	.52204	41.621	.25177	4.9644	.24365	.60722	.18517	7.6439
#2	.51965	40.900	.24967	4.9597	.25336	.61129	.19249	7.6237
#3	.52125	40.985	.24970	4.9549	.25223	.60580	.19497	7.6359
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606080803S Acquired: 6/23/2016 20:18:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-07

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50663	.69353	.50821	.24979	.50502	.50112	.56844
Stddev	.00206	.00495	.00984	.00132	.00161	.00059	.24983
%RSD	.40682	.71328	1.9372	.52651	.31873	.11783	43.949
#1	.50895	.69918	.50790	.24829	.50330	.50110	.76022
#2	.50591	.69001	.49852	.25075	.50528	.50054	.28593
#3	.50502	.69139	.51821	.25034	.50649	.50172	.65917
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6625.1	41184.	2141.2
Stddev	43.1	220.	34.4
%RSD	.65093	.53421	1.6047
#1	6575.6	40948.	2101.5
#2	6654.8	41222.	2162.6
#3	6644.9	41383.	2159.3

Sample Name: L1606080804SD Acquired: 6/23/2016 20:21:47 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20417	5.0253	.20216	1.0213	.52389	.02512	58.507	.02515
Stddev	.00061	.0013	.00476	.0019	.00456	.00016	.379	.00007
%RSD	.29699	.02608	2.3559	.18158	.87021	.62066	.64724	.26301
#1	.20348	5.0268	.19939	1.0192	.52806	.02496	58.865	.02518
#2	.20461	5.0245	.20766	1.0224	.52461	.02527	58.546	.02507
#3	.20443	5.0246	.19943	1.0224	.51902	.02514	58.111	.02519
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10013	.25277	.25192	2.3449	27.111	.52740	20.302	.37943
Stddev	.00006	.00076	.00111	.0178	.231	.00552	.216	.00380
%RSD	.06407	.29922	.44124	.75833	.85294	1.0468	1.0620	1.0007
#1	.10006	.25267	.25295	2.3582	27.342	.53203	20.543	.38169
#2	.10018	.25357	.25074	2.3247	27.111	.52888	20.233	.38156
#3	.10015	.25207	.25208	2.3517	26.880	.52129	20.129	.37505
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.52376	41.862	.25037	4.9891	.25282	.61284	.19379	7.7402
Stddev	.00119	.339	.00131	.0038	.00458	.00414	.01024	.0102
%RSD	.22762	.81017	.52333	.07550	1.8112	.67508	5.2847	.13142
#1	.52513	42.192	.25099	4.9856	.25772	.61266	.19670	7.7511
#2	.52310	41.879	.25125	4.9931	.24866	.60879	.20226	7.7310
#3	.52304	41.514	.24886	4.9887	.25208	.61705	.18241	7.7384
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606080804SD Acquired: 6/23/2016 20:21:47 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-08

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50881**	**.70095**	**.50511**	**.25346**	**.50822**	**.50328**	**.36568**
Stddev	.00118	.00336	.00758	.00055	.00107	.00052	.09051
%RSD	.23262	.47943	1.5010	.21780	.21111	.10273	24.750
#1	.50993	.70351	.51202	.25388	.50727	.50277	.45775
#2	.50891	.70221	.50631	.25366	.50938	.50380	.36245
#3	.50757	.69715	.49700	.25283	.50801	.50326	.27683
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6648.3**	**41256.**	**2167.1**
Stddev	37.5	212.	35.6
%RSD	.56386	.51441	1.6437
#1	6687.1	41426.	2128.5
#2	6645.3	41018.	2174.1
#3	6612.3	41324.	2198.7

Sample Name: L1606082202 Acquired: 6/23/2016 20:25:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00152	.02026	-.00003	.10248	.15200	.00003	6.5973	.00005
Stddev	.00174	.00884	.00264	.00097	.00250	.00003	.1001	.00016
%RSD	114.35	43.656	8042.8	.94662	1.6469	112.93	1.5178	320.68
#1	-.00041	.02197	.00037	.10136	.15374	.00007	6.6625	.00019
#2	.00200	.02813	-.00285	.10296	.15311	.00001	6.6474	.00008
#3	.00298	.01069	.00238	.10312	.14913	.00001	6.4820	-.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	-.00052	.00142	2.2172	1.7251	.02347	4.6721	.08008
Stddev	.00019	.00160	.00145	.0203	.1370	.00447	.1537	.00373
%RSD	39.028	307.12	102.30	.91315	7.9439	19.041	3.2904	4.6534
#1	.00065	-.00214	.00215	2.1938	1.8809	.02224	4.7545	.08412
#2	.00052	.00105	-.00025	2.2287	1.6712	.01975	4.7670	.07933
#3	.00028	-.00047	.00236	2.2290	1.6232	.02843	4.4947	.07679
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00084	87.069	.00079	.09610	.00066	.00367	-.00338	12.803
Stddev	.00076	.850	.00109	.00417	.00060	.00400	.00476	.010
%RSD	90.058	.97655	138.31	4.3415	91.329	109.21	141.00	.07526
#1	-.00000	87.844	-.00046	.09129	-.00003	.00288	-.00767	12.795
#2	.00146	87.204	.00156	.09839	.00108	.00011	.00175	12.801
#3	.00107	86.159	.00126	.09864	.00092	.00800	-.00421	12.814
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606082202 Acquired: 6/23/2016 20:25:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00121	.22077	-.00522	-.00363	.00085	.00236	.22639
Stddev	.00050	.00200	.00719	.00112	.00069	.00021	.20548
%RSD	41.542	.90585	137.85	30.833	80.735	8.9617	90.766
#1	-.00156	.22257	-.00824	-.00440	.00009	.00222	.33820
#2	-.00063	.22113	-.01041	-.00235	.00143	.00261	.35171
#3	-.00143	.21862	.00299	-.00415	.00104	.00226	-.01076
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6664.8	41248.	2126.9
Stddev	26.9	323.	26.2
%RSD	.40377	.78403	1.2328
#1	6695.8	40876.	2101.3
#2	6647.4	41460.	2125.7
#3	6651.2	41408.	2153.7

Sample Name: L1606082203 Acquired: 6/23/2016 20:29:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00153	.05110	.00192	.03595	.79888	.00004	32.443	.00019
Stddev	.00257	.00394	.00088	.00096	.00490	.00006	.223	.00012
%RSD	168.19	7.7097	45.736	2.6764	.61347	156.15	.68852	63.266
#1	.00450	.04867	.00097	.03585	.80394	.00006	32.700	.00006
#2	.00008	.05565	.00210	.03696	.79416	-.00003	32.293	.00030
#3	.00001	.04899	.00269	.03505	.79852	.00008	32.336	.00020
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00307	.00707	.00126	5.3792	51.825	.10776	15.276	.18167
Stddev	.00032	.00094	.00080	.0419	.270	.00194	.142	.00297
%RSD	10.345	13.245	64.108	.77858	.52117	1.8035	.92803	1.6345
#1	.00299	.00617	.00218	5.4159	52.078	.10909	15.426	.17840
#2	.00280	.00804	.00087	5.3881	51.541	.10866	15.144	.18240
#3	.00342	.00701	.00071	5.3336	51.857	.10553	15.259	.18420
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00008	93.634	.00315	-.00252	.00643	.00340	-.00637	11.872
Stddev	.00025	.401	.00140	.00666	.00138	.00643	.00382	.014
%RSD	298.72	.42875	44.327	264.34	21.428	189.23	59.894	.11647
#1	-.00020	94.085	.00392	-.00701	.00774	.00248	-.00439	11.857
#2	.00021	93.316	.00154	.00513	.00656	.01023	-.00396	11.885
#3	.00024	93.502	.00400	-.00568	.00499	-.00252	-.01077	11.874
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606082203 Acquired: 6/23/2016 20:29:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00017**	**.82666**	**-.00447**	**-.00598**	**-.00056**	**.00596**	**.27009**
Stddev	.00125	.01637	.00340	.00085	.00033	.00013	.03306
%RSD	751.65	1.9807	76.005	14.225	59.976	2.2544	12.242
#1	.00127	.84555	-.00552	-.00599	-.00043	.00586	.28137
#2	-.00101	.81782	-.00722	-.00682	-.00030	.00611	.29603
#3	-.00076	.81660	-.00067	-.00512	-.00093	.00591	.23286
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6576.7**	**40904.**	**2132.1**
Stddev	24.6	214.	16.8
%RSD	.37432	.52260	.78563
#1	6552.2	40790.	2112.8
#2	6576.5	40771.	2141.0
#3	6601.4	41150.	2142.6

Sample Name: L1606082302 Acquired: 6/23/2016 20:33:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00184	.02080	.00098	.02231	7.8783	.00017	255.75
Stddev	.00203	.01187	.00435	.00335	.0432	.00007	1.49
%RSD	110.13	57.087	445.14	15.024	.54831	41.236	.58087
#1	-.00150	.00710	.00584	.02555	7.9220	.00012	256.99
#2	-.00401	.02815	-.00256	.01885	7.8356	.00014	254.11
#3	-.00000	.02715	-.00034	.02253	7.8772	.00025	256.16
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00094	.05342	.00242	.00486	4.7469	3.4904	.13833
Stddev	.00052	.00070	.00059	.00069	.0046	.0030	.00615
%RSD	55.632	1.3175	24.430	14.175	.09644	.08679	4.4491
#1	.00040	.05380	.00272	.00415	4.7522	3.4870	.14015
#2	.00098	.05261	.00174	.00553	4.7443	3.4928	.13147
#3	.00144	.05385	.00281	.00490	4.7442	3.4915	.14336
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	198.54	.84118	-.00019	F 709.80	.01731	.04470	.00329
Stddev	1.62	.00623	.00016	4.09	.00125	.01000	.00583
%RSD	.81609	.74094	85.381	.57589	7.2164	22.367	177.43
#1	200.41	.84837	-.00014	712.90	.01830	.03650	.00825
#2	197.48	.83796	-.00006	705.16	.01772	.04176	.00474
#3	197.74	.83722	-.00036	711.32	.01590	.05584	-.00313
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Sample Name: L1606082302 Acquired: 6/23/2016 20:33:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00699**	**-.00211**	**18.432**	**-.00053**	**5.9869**	F **-.04100**	**-.00343**
Stddev	.00326	.00362	.021	.00131	.0347	.00868	.00076
%RSD	46.731	171.57	.11640	247.46	.57940	21.163	22.234
#1	.00892	.00180	18.414	-.00131	6.0214	-.04734	-.00288
#2	.00882	-.00536	18.426	-.00125	5.9520	-.03111	-.00311
#3	.00322	-.00277	18.456	.00098	5.9872	-.04457	-.00430
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00262**	**.00921**	F **-.07344**
Stddev	.00140	.00009	.18020
%RSD	53.460	.99471	245.36
#1	-.00401	.00912	.10768
#2	-.00261	.00930	-.07530
#3	-.00122	.00920	-.25270
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6211.9**	**37943.**	**2065.9**
Stddev	10.7	171.	36.3
%RSD	.17281	.45038	1.7553
#1	6203.2	38131.	2024.8
#2	6223.9	37902.	2079.8
#3	6208.6	37796.	2093.2

Sample Name: L1606082303 Acquired: 6/23/2016 20:37:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00036	.09100	.00296	.01304	4.1939	-.00007	203.18
Stddev	.00140	.00531	.00327	.00376	.0297	.00006	1.67
%RSD	388.51	5.8358	110.60	28.801	.70801	88.134	.82056
#1	-.00165	.09683	.00075	.00892	4.2189	-.00004	204.69
#2	.00112	.08975	.00671	.01627	4.1611	-.00014	201.39
#3	-.00055	.08643	.00141	.01393	4.2018	-.00002	203.46
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00062	.00605	.00148	.00325	.50161	1.7415	.06042
Stddev	.00049	.00063	.00113	.00080	.03478	.1908	.00296
%RSD	79.462	10.371	76.602	24.650	6.9332	10.958	4.9058
#1	.00115	.00647	.00038	.00343	.46408	1.6330	.05783
#2	.00054	.00533	.00142	.00237	.53275	1.6297	.06365
#3	.00017	.00634	.00265	.00394	.50800	1.9619	.05979
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	118.15	2.3989	.00023	F 640.12	.00507	.01016	.00289
Stddev	.81	.0242	.00018	10.00	.00123	.00431	.00149
%RSD	.68254	1.0089	76.880	1.5615	24.328	42.407	51.612
#1	118.94	2.4148	.00038	650.69	.00487	.00561	.00152
#2	117.33	2.3711	.00027	630.82	.00639	.01069	.00267
#3	118.17	2.4110	.00003	638.86	.00395	.01418	.00448
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Sample Name: L1606082303 Acquired: 6/23/2016 20:37:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00012**	**-.00944**	**13.225**	**-.00084**	**3.8421**	F **-.03122**	**-.00075**
Stddev	.00335	.00541	.023	.00156	.0312	.00884	.00263
%RSD	2724.2	57.378	.17522	185.97	.81266	28.325	351.62
#1	.00362	-.00548	13.247	.00011	3.8716	-.02289	.00226
#2	-.00306	-.00723	13.225	-.00264	3.8094	-.04050	-.00263
#3	-.00019	-.01560	13.201	.00001	3.8453	-.03028	-.00187
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00215**	**.00570**	**.39111**
Stddev	.00161	.00012	.18787
%RSD	74.559	2.0758	48.034
#1	-.00169	.00568	.30759
#2	-.00394	.00583	.60626
#3	-.00083	.00559	.25948
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6278.4**	**38565.**	**2078.5**
Stddev	22.0	175.	41.0
%RSD	.35101	.45432	1.9726
#1	6265.3	38533.	2033.9
#2	6266.0	38754.	2114.5
#3	6303.8	38408.	2087.2

Sample Name: L1606082304 Acquired: 6/23/2016 20:41:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00041**	**.00982**	**.00065**	**.03078**	**1.3316**	**-.00001**	**46.431**
Stddev	.00324	.01473	.00404	.00488	.0087	.00012	.330
%RSD	781.99	150.01	619.18	15.855	.65651	2204.3	.71076
#1	.00226	.02597	.00499	.03065	1.3415	-.00014	46.804
#2	.00231	.00636	-.00301	.03572	1.3252	.00009	46.176
#3	-.00333	-.00287	-.00002	.02596	1.3280	.00003	46.315
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00066**	**-.00016**	F **-.00347**	**.00119**	**84.345**	**3.2107**	**.04191**
Stddev	.00021	.00040	.00039	.00149	.339	.1063	.00251
%RSD	31.873	248.01	11.380	125.50	.40248	3.3107	5.9779
#1	.00060	-.00031	-.00306	.00290	84.661	3.3331	.04445
#2	.00048	-.00046	-.00385	.00044	84.387	3.1411	.03944
#3	.00089	.00029	-.00349	.00022	83.986	3.1580	.04185
Check ?	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit			36.000				
Low Limit			-.00300				

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**29.094**	**1.4706**	**-.00067**	**141.52**	**.00104**	**.07644**	**.00312**
Stddev	.106	.0113	.00040	1.02	.00036	.00458	.00219
%RSD	.36352	.76758	59.671	.72353	34.222	5.9922	69.972
#1	29.216	1.4831	-.00079	142.69	.00136	.08096	.00145
#2	29.030	1.4675	-.00100	141.08	.00110	.07654	.00233
#3	29.037	1.4612	-.00023	140.78	.00066	.07180	.00560
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606082304 Acquired: 6/23/2016 20:41:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**-.00444**	**16.946**	**-.00016**	**1.0594**	**.01618**	**-.00009**
Stddev	.00138	.00283	.014	.00045	.0066	.00613	.00273
%RSD	1084.7	63.725	.08315	282.51	.62739	37.902	3068.2
#1	-.00159	-.00414	16.934	.00035	1.0669	.01325	.00044
#2	.00005	-.00741	16.943	-.00032	1.0545	.02323	.00234
#3	.00116	-.00177	16.961	-.00051	1.0567	.01207	-.00305
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00431**	**.01323**	F **-.28757**
Stddev	.00070	.00016	.06453
%RSD	16.230	1.2047	22.439
#1	-.00354	.01326	-.33446
#2	-.00489	.01337	-.31427
#3	-.00452	.01306	-.21398
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6548.1**	**40764.**	**2104.8**
Stddev	18.7	118.	30.0
%RSD	.28493	.28837	1.4234
#1	6562.8	40898.	2079.6
#2	6554.4	40678.	2096.8
#3	6527.1	40717.	2137.9

Sample Name: L1606088302 Acquired: 6/23/2016 20:45:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00335	.84495	.01492	.13705	.35776	-.00016	61.342	.00073
Stddev	.00192	.00536	.00123	.00168	.00649	.00001	.914	.00037
%RSD	57.476	.63420	8.2502	1.2235	1.8144	6.0991	1.4894	50.484
#1	.00126	.85090	.01399	.13791	.36483	-.00016	62.379	.00100
#2	.00372	.84051	.01632	.13512	.35638	-.00017	60.987	.00089
#3	.00506	.84344	.01445	.13813	.35207	-.00015	60.658	.00031
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00790	.00132	.00561	5.6286	11.735	.00962	26.311	2.7437
Stddev	.00012	.00055	.00131	.0909	.087	.00112	.394	.0396
%RSD	1.4821	41.328	23.360	1.6145	.74143	11.624	1.4976	1.4443
#1	.00790	.00105	.00615	5.6698	11.835	.00907	26.689	2.7892
#2	.00779	.00096	.00657	5.6916	11.681	.00887	26.343	2.7252
#3	.00802	.00195	.00412	5.5244	11.688	.01090	25.902	2.7167
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01337	55.014	.02594	.07391	.01768	.00363	-.00148	7.7740
Stddev	.00021	.737	.00058	.00987	.00156	.00456	.00201	.0095
%RSD	1.5352	1.3397	2.2195	13.356	8.8413	125.65	135.81	.12214
#1	.01329	55.834	.02652	.06415	.01943	.00884	.00061	7.7638
#2	.01322	54.801	.02594	.08389	.01641	.00164	-.00340	7.7755
#3	.01361	54.406	.02537	.07370	.01720	.00040	-.00164	7.7826
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606088302 Acquired: 6/23/2016 20:45:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00090	.36986	.01540	-.00118	.00118	.12530	1.4316
Stddev	.00025	.00460	.01415	.00177	.00035	.00030	.3050
%RSD	27.990	1.2450	91.875	150.26	29.961	.24276	21.302
#1	.00061	.37482	.00658	-.00313	.00110	.12559	1.0819
#2	.00104	.36904	.00789	.00033	.00156	.12533	1.6421
#3	.00105	.36572	.03171	-.00073	.00087	.12498	1.5707
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	7921.9	49318.	2693.6
Stddev	12.3	123.	25.0
%RSD	.15529	.24979	.92714
#1	7908.5	49427.	2667.6
#2	7924.8	49342.	2717.4
#3	7932.6	49184.	2695.7

Sample Name: CCV Acquired: 6/23/2016 20:49:10 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39590	9.9110	.39526	.48909	.98565	.04996	9.8451
Stddev	.00120	.0339	.00042	.00457	.02803	.00042	.3357
%RSD	.30353	.34216	.10738	.93432	2.8436	.84013	3.4103
#1	.39483	9.8747	.39503	.48786	.95368	.04993	9.4641
#2	.39720	9.9418	.39501	.48526	.99729	.04955	9.9733
#3	.39566	9.9167	.39575	.49415	1.0060	.05039	10.098
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05017	.20113	.49806	.50390	3.9193	49.281	.99930
Stddev	.00013	.00026	.00295	.00102	.1281	1.426	.02793
%RSD	.24979	.12923	.59274	.20262	3.2691	2.8935	2.7947
#1	.05003	.20121	.49990	.50316	3.7715	47.645	.96745
#2	.05025	.20083	.49465	.50347	3.9883	49.937	1.0109
#3	.05024	.20133	.49963	.50506	3.9981	50.261	1.0196
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.8643	.49902	1.0085	49.184	.49977	9.8773	.49959
Stddev	.3185	.01772	.0043	1.350	.00010	.0178	.00372
%RSD	3.2291	3.5503	.42423	2.7456	.01960	.18066	.74560
#1	9.4975	.47954	1.0116	47.634	.49976	9.8611	.50047
#2	10.025	.50333	1.0103	49.812	.49987	9.8744	.49550
#3	10.071	.51418	1.0036	50.106	.49968	9.8964	.50280
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 20:49:10 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2048**	**.40490**	**5.0297**	**1.0004**	**.98725**	**.99313**	**.50297**
Stddev	.0042	.00796	.0039	.0034	.02832	.03209	.00363
%RSD	.34965	1.9650	.07741	.34428	2.8681	3.2307	.72121
#1	1.2087	.41298	5.0259	.99890	.95492	.95803	.50669
#2	1.2003	.40465	5.0296	.99795	.99923	1.0004	.49944
#3	1.2055	.39708	5.0337	1.0043	1.0076	1.0209	.50278
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.99558**	**1.0009**	F **.02289**
Stddev	.00580	.0022	.25371
%RSD	.58246	.21707	1108.2
#1	.99450	.99842	.14548
#2	.99040	1.0017	-.26883
#3	1.0018	1.0025	.19203
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%
Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6660.9**	**41801.**	**2169.0**
Stddev	32.3	28.	115.6
%RSD	.48534	.06597	5.3274
#1	6698.0	41774.	2302.5
#2	6639.4	41798.	2102.0
#3	6645.2	41829.	2102.6

Sample Name: CCB Acquired: 6/23/2016 20:52:47 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00186	-.00093	.00316	.00230	.00074	.00006	.02069
Stddev	.00123	.00851	.00342	.00122	.00124	.00008	.05541
%RSD	66.210	911.24	108.19	52.892	166.55	142.19	267.79
#1	.00059	-.01042	.00649	.00167	.00054	-.00003	-.00744
#2	.00195	.00160	.00333	.00153	-.00038	.00006	-.01500
#3	.00305	.00602	-.00034	.00370	.00206	.00014	.08452
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00002	.00039	.00115	.00076	-.01535	.06062	-.00386
Stddev	.00025	.00015	.00083	.00156	.01904	.09913	.00805
%RSD	1273.3	38.748	71.888	205.02	124.01	163.52	208.55
#1	.00015	.00037	.00104	.00077	-.03627	-.05039	.00540
#2	-.00027	.00056	.00039	.00232	.00098	.09196	-.00780
#3	.00017	.00026	.00203	-.00080	-.01077	.14030	-.00918
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01451	-.00216	.00311	.14026	-.00019	.00962	.00081
Stddev	.08285	.00307	.00109	.02230	.00100	.00497	.00347
%RSD	570.85	141.96	34.905	15.899	521.75	51.686	430.66
#1	-.05222	-.00061	.00206	.16585	-.00076	.01507	.00163
#2	.10725	-.00570	.00305	.12497	-.00078	.00533	-.00300
#3	-.01149	-.00018	.00423	.12997	.00096	.00846	.00380
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 20:52:47 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00098**	**-.00415**	**.00357**	**.00020**	**.00045**	**-.00310**	**-.00314**
Stddev	.00119	.00276	.00144	.00034	.00013	.00169	.00399
%RSD	121.00	66.464	40.393	168.69	30.219	54.451	127.02
#1	-.00049	-.00098	.00477	.00034	.00060	-.00134	-.00058
#2	-.00234	-.00542	.00396	-.00019	.00034	-.00325	-.00773
#3	-.00012	-.00603	.00197	.00045	.00040	-.00471	-.00111
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00114**	**.00020**	F **.30264**
Stddev	.00114	.00013	.32767
%RSD	99.237	64.632	108.27
#1	-.00245	.00019	.39170
#2	-.00061	.00033	-.06036
#3	-.00037	.00008	.57657
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000
Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6711.0**	**41762.**	**2107.6**
Stddev	39.0	171.	14.5
%RSD	.58138	.41024	.68723
#1	6710.2	41729.	2123.6
#2	6672.4	41947.	2095.3
#3	6750.5	41609.	2103.9

Sample Name: L1606088401 Acquired: 6/23/2016 20:56:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00138**	**.01701**	**.00493**	**.04852**	**.00618**	**-.00007**	**12.628**	**.00247**
Stddev	.00005	.00414	.00116	.00188	.00077	.00012	.147	.00032
%RSD	3.9151	24.333	23.511	3.8697	12.414	161.67	1.1647	12.874
#1	-.00141	.01981	.00570	.04698	.00662	.00006	12.792	.00280
#2	-.00141	.01895	.00549	.04798	.00529	-.00016	12.508	.00217
#3	-.00131	.01225	.00360	.05061	.00662	-.00013	12.584	.00245
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.33204**	**-.00005**	**.12713**	**.00004**	**.29521**	**1.5328**	**3.2466**	**.01448**
Stddev	.00034	.00102	.00048	.01339	.06956	.0143	.1805	.00109
%RSD	.10335	2250.3	.37814	32463.	23.564	.93016	5.5601	7.5480
#1	.33185	.00032	.12750	.00310	.34102	1.5490	3.4388	.01457
#2	.33183	.00074	.12659	-.01461	.32944	1.5271	3.0806	.01335
#3	.33243	-.00120	.12731	.01164	.21517	1.5222	3.2204	.01553
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00093**	**152.98**	**10.187**	**.09320**	**-.00036**	**-.00902**	**-.00316**	**.51224**
Stddev	.00032	1.78	.001	.01280	.00229	.00381	.00857	.00180
%RSD	34.564	1.1612	.01164	13.735	636.44	42.238	271.43	.35176
#1	.00071	155.02	10.189	.07854	-.00010	-.00788	-.00504	.51429
#2	.00079	151.86	10.188	.09884	-.00276	-.00591	.00620	.51148
#3	.00130	152.04	10.186	.10221	.00179	-.01327	-.01063	.51094
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606088401 Acquired: 6/23/2016 20:56:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00055	.06978	-.00044	-.00366	-.00050	.00520	.35447
Stddev	.00021	.00053	.00665	.00240	.00048	.00018	.13127
%RSD	37.225	.76020	1503.3	65.540	95.032	3.3899	37.033
#1	.00077	.07037	-.00450	-.00503	-.00086	.00501	.20475
#2	.00053	.06936	-.00405	-.00089	-.00069	.00524	.40886
#3	.00036	.06960	.00723	-.00506	.00004	.00536	.44980
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6642.8	41062.	2131.9
Stddev	24.2	57.	30.4
%RSD	.36471	.13951	1.4260
#1	6639.2	41009.	2097.0
#2	6620.6	41122.	2152.2
#3	6668.6	41053.	2146.6

Sample Name: L1606089601 Acquired: 6/23/2016 21:00:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00276	.06592	.00541	.02208	.05123	-.00005	4.5311	.00038
Stddev	.00137	.00695	.00371	.00073	.00110	.00012	.0222	.00020
%RSD	49.858	10.542	68.509	3.3035	2.1551	257.77	.49075	51.621
#1	.00357	.06862	.00418	.02271	.05250	.00006	4.5567	.00056
#2	.00353	.05803	.00958	.02226	.05063	-.00017	4.5198	.00017
#3	.00117	.07112	.00247	.02128	.05055	-.00003	4.5168	.00042
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.06787	-.00161	.00146	42.996	.90153	-.00084	5.1151	6.0536
Stddev	.00044	.00070	.00153	.463	.08378	.00102	.1337	.0524
%RSD	.64651	43.319	104.57	1.0779	9.2930	121.94	2.6141	.86472
#1	.06763	-.00220	.00017	43.521	.97406	.00034	5.2374	6.1132
#2	.06761	-.00084	.00314	42.825	.80982	-.00134	4.9723	6.0323
#3	.06838	-.00179	.00107	42.642	.92069	-.00151	5.1355	6.0152
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00036	19.350	.00357	-.00356	.00111	.00096	-.00499	4.5346
Stddev	.00020	.125	.00141	.00838	.00156	.00212	.00300	.0153
%RSD	55.732	.64532	39.385	235.57	139.61	221.63	60.052	.33782
#1	-.00054	19.492	.00220	-.01302	-.00064	.00306	-.00344	4.5188
#2	-.00040	19.300	.00350	.00291	.00231	-.00118	-.00845	4.5355
#3	-.00014	19.259	.00500	-.00056	.00168	.00099	-.00309	4.5494
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089601 Acquired: 6/23/2016 21:00:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00060**	**.03780**	**.00811**	**-.00381**	**-.00115**	**.00560**	**.10083**
Stddev	.00066	.00049	.00739	.00382	.00028	.00007	.23956
%RSD	110.76	1.2963	91.111	100.26	24.644	1.2642	237.60
#1	.00016	.03802	.00291	-.00026	-.00147	.00553	-.17268
#2	-.00087	.03815	.00485	-.00331	-.00104	.00559	.20174
#3	-.00109	.03724	.01658	-.00784	-.00094	.00567	.27342
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6699.7**	**41605.**	**2147.2**
Stddev	6.0	395.	20.8
%RSD	.08901	.94822	.96897
#1	6705.8	41196.	2129.6
#2	6699.3	41634.	2141.8
#3	6693.9	41984.	2170.2

Sample Name: L1606089602 Acquired: 6/23/2016 21:04:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00115	.04083	.00060	.00715	.00811	-.00006	.41120	.00010
Stddev	.00092	.00909	.00372	.00237	.00130	.00006	.07420	.00010
%RSD	79.895	22.273	617.42	33.111	16.069	103.37	18.045	95.174
#1	.00172	.03036	-.00348	.00464	.00961	-.00013	.44439	.00004
#2	.00163	.04672	.00380	.00935	.00726	-.00004	.46300	.00021
#3	.00009	.04541	.00149	.00745	.00746	-.00001	.32619	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00002	.00085	.00131	.07554	1.8076	-.00047	.25019	.00154
Stddev	.00047	.00183	.00071	.02271	.0472	.00199	.07424	.00073
%RSD	2125.6	214.06	54.320	30.062	2.6112	424.18	29.673	47.627
#1	-.00052	.00002	.00060	.10056	1.7705	.00136	.32911	.00152
#2	.00004	.00295	.00130	.05624	1.7915	-.00018	.18174	.00229
#3	.00042	-.00040	.00202	.06981	1.8607	-.00259	.23972	.00082
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00001	8.3070	.00081	-.00279	.00077	-.00363	-.00191	9.3341
Stddev	.00042	.0355	.00152	.00730	.00368	.00212	.00752	.0120
%RSD	4219.9	.42734	186.83	261.51	478.27	58.376	393.78	.12811
#1	-.00026	8.3438	.00114	-.00340	.00064	-.00604	.00678	9.3471
#2	.00047	8.2730	-.00084	.00479	-.00284	-.00208	-.00623	9.3318
#3	-.00024	8.3042	.00214	-.00976	.00451	-.00276	-.00628	9.3235
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089602 Acquired: 6/23/2016 21:04:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00104**	**.00271**	**.00141**	**-.00441**	**.00055**	**.00294**	**.09481**
Stddev	.00076	.00065	.00707	.00096	.00059	.00032	.12117
%RSD	73.387	23.817	499.85	21.749	106.89	10.896	127.81
#1	-.00017	.00268	.00405	-.00333	.00004	.00305	.13413
#2	-.00159	.00337	.00678	-.00473	.00120	.00257	-.04114
#3	-.00135	.00208	-.00659	-.00517	.00043	.00318	.19143
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6686.8**	**41808.**	**2134.5**
Stddev	18.1	213.	20.0
%RSD	.27005	.50969	.93843
#1	6699.2	41631.	2111.6
#2	6666.1	41748.	2148.9
#3	6695.1	42044.	2142.9

Sample Name: L1606089603 Acquired: 6/23/2016 21:08:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00102	.11801	.00474	.00963	.06515	.00009	2.8598	-.00002
Stddev	.00304	.00302	.00162	.00102	.00099	.00007	.0117	.00049
%RSD	298.48	2.5577	34.279	10.645	1.5207	83.968	.40938	2310.2
#1	.00427	.12112	.00606	.01075	.06530	.00010	2.8502	-.00015
#2	.00053	.11781	.00292	.00939	.06607	.00015	2.8729	-.00043
#3	-.00175	.11509	.00524	.00874	.06410	.00001	2.8564	.00052
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00155	-.00024	.00127	2.7358	3.5730	.00040	1.7276	.02916
Stddev	.00022	.00132	.00125	.0173	.1870	.00308	.0541	.00143
%RSD	13.914	551.57	98.352	.63220	5.2335	762.16	3.1323	4.9029
#1	.00170	-.00168	.00263	2.7557	3.7847	.00309	1.6947	.02869
#2	.00164	.00005	.00099	2.7247	3.4302	-.00295	1.7900	.03077
#3	.00130	.00091	.00018	2.7268	3.5042	.00107	1.6980	.02803
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00031	7.6284	.00235	.00047	-.00082	.00245	-.00905	5.4943
Stddev	.00032	.0301	.00079	.01031	.00433	.00306	.00148	.0133
%RSD	102.08	.39457	33.547	2191.2	526.63	125.12	16.371	.24152
#1	-.00063	7.6521	.00324	.00549	-.00579	.00459	-.00977	5.4890
#2	-.00033	7.6386	.00175	-.01139	.00215	-.00106	-.01004	5.4845
#3	.00001	7.5946	.00205	.00731	.00118	.00382	-.00735	5.5094
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089603 Acquired: 6/23/2016 21:08:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00012	.02218	.00164	-.00426	-.00004	.00835	.14767
Stddev	.00040	.00013	.00288	.00161	.00009	.00018	.44819
%RSD	344.86	.60531	175.60	37.721	191.36	2.1138	303.52
#1	.00047	.02205	-.00008	-.00574	-.00012	.00836	-.32337
#2	-.00031	.02232	.00496	-.00255	.00005	.00853	.19752
#3	.00019	.02218	.00004	-.00448	-.00006	.00817	.56885
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6682.9	41728.	2128.8
Stddev	17.5	288.	26.5
%RSD	.26233	.68938	1.2452
#1	6675.1	41447.	2098.3
#2	6670.6	42022.	2146.0
#3	6703.0	41715.	2142.2

Sample Name: L1606089604 Acquired: 6/23/2016 21:12:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00189**	**.07787**	**.00274**	**.02534**	**.04671**	**-.00007**	**4.6127**
Stddev	.00185	.01361	.00115	.00174	.00203	.00009	.0084
%RSD	97.576	17.475	41.979	6.8483	4.3441	132.30	.18280
#1	.00394	.07337	.00237	.02334	.04637	.00002	4.6224
#2	.00037	.09315	.00403	.02650	.04487	-.00016	4.6067
#3	.00136	.06707	.00182	.02616	.04888	-.00006	4.6091
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00046**	**.07375**	**-.00139**	**.00045**	**44.265**	**.94195**	**-.00388**
Stddev	.00010	.00022	.00095	.00149	.255	.04262	.00409
%RSD	21.656	.29660	68.674	332.55	.57666	4.5243	105.62
#1	.00034	.07398	-.00166	-.00009	44.524	.94561	-.00857
#2	.00051	.07354	-.00218	-.00070	44.257	.89762	-.00204
#3	.00052	.07372	-.00033	.00214	44.014	.98262	-.00102
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**6.3790**	**6.1215**	**-.00097**	**19.253**	**.00495**	**-.00232**	**.00060**
Stddev	.0237	.0305	.00040	.056	.00026	.00993	.00262
%RSD	.37160	.49866	40.991	.29188	5.2685	429.04	440.43
#1	6.3989	6.1509	-.00053	19.318	.00524	-.00316	-.00241
#2	6.3528	6.1237	-.00131	19.216	.00473	-.01180	.00179
#3	6.3853	6.0899	-.00108	19.227	.00489	.00801	.00241
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606089604 Acquired: 6/23/2016 21:12:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00465**	**-.00637**	**4.5161**	**-.00039**	**.04033**	**.00878**	**-.00291**
Stddev	.00319	.00325	.0073	.00021	.00012	.01173	.00069
%RSD	68.469	51.078	.16173	52.620	.29003	133.62	23.851
#1	.00106	-.00941	4.5223	-.00056	.04037	.01678	-.00220
#2	.00578	-.00675	4.5081	-.00016	.04019	.01426	-.00359
#3	.00712	-.00294	4.5179	-.00046	.04041	-.00469	-.00294
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00260**	**.00552**	F **-.09796**
Stddev	.00057	.00025	.21094
%RSD	21.873	4.5503	215.32
#1	-.00325	.00580	.14398
#2	-.00238	.00531	-.24325
#3	-.00218	.00544	-.19462
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6689.8**	**41666.**	**2101.2**
Stddev	9.9	402.	24.3
%RSD	.14842	.96487	1.1576
#1	6695.2	41372.	2073.4
#2	6678.4	42124.	2111.3
#3	6695.9	41503.	2118.8

Sample Name: L1606089605 Acquired: 6/23/2016 21:16:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00006	.04830	-.00145	.00590	.06272	.00007	2.6789	.00041
Stddev	.00176	.01001	.00427	.00202	.00179	.00010	.0329	.00039
%RSD	2993.4	20.730	295.15	34.254	2.8603	144.40	1.2294	96.299
#1	-.00190	.05947	.00328	.00605	.06424	-.00005	2.7135	.00011
#2	.00153	.04527	-.00502	.00784	.06318	.00013	2.6479	.00085
#3	.00055	.04014	-.00260	.00381	.06074	.00012	2.6754	.00025
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00157	-.00035	.00184	.89250	3.4230	-.00452	1.6804	.01911
Stddev	.00028	.00023	.00080	.02843	.0162	.00117	.0608	.00353
%RSD	17.515	67.269	43.308	3.1854	.47175	25.991	3.6163	18.451
#1	.00131	-.00060	.00162	.86545	3.4125	-.00336	1.6187	.02198
#2	.00155	-.00014	.00273	.92213	3.4148	-.00448	1.7402	.01517
#3	.00186	-.00031	.00117	.88990	3.4416	-.00571	1.6824	.02016
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00024	7.0650	.00185	-.00101	-.00247	.00326	-.00672	5.2164
Stddev	.00018	.0411	.00050	.00181	.00243	.00063	.00650	.0100
%RSD	75.478	.58180	27.196	179.64	98.576	19.406	96.645	.19223
#1	-.00033	7.1107	.00239	-.00299	-.00232	.00293	.00077	5.2089
#2	-.00037	7.0530	.00139	.00057	-.00497	.00286	-.01006	5.2125
#3	-.00003	7.0312	.00176	-.00061	-.00011	.00399	-.01087	5.2278
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089605 Acquired: 6/23/2016 21:16:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00041	.02070	-.00116	-.00495	-.00050	.00647	.22434
Stddev	.00061	.00025	.00062	.00270	.00039	.00006	.14133
%RSD	147.06	1.2106	53.841	54.602	79.342	1.0037	62.996
#1	-.00061	.02054	-.00094	-.00697	-.00093	.00640	.33918
#2	.00027	.02099	-.00186	-.00188	-.00039	.00649	.26733
#3	-.00089	.02057	-.00067	-.00601	-.00017	.00652	.06651
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6679.9	41653.	2126.5
Stddev	8.5	193.	24.2
%RSD	.12769	.46319	1.1397
#1	6686.1	41709.	2098.6
#2	6670.2	41811.	2142.5
#3	6683.5	41438.	2138.4

Sample Name: L1606089607 Acquired: 6/23/2016 21:20:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00019**	**.00633**	**.00159**	**.00141**	**-.00044**	**-.00001**	**.07933**
Stddev	.00195	.00666	.00483	.00232	.00080	.00015	.08022
%RSD	1049.3	105.18	304.39	163.83	179.45	1810.1	101.12
#1	-.00175	.01203	-.00042	.00047	-.00070	-.00018	.13178
#2	.00017	.00795	.00710	-.00028	-.00108	.00005	-.01301
#3	.00214	-.00099	-.00192	.00405	.00045	.00010	.11922
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00011**	**.00023**	**.00018**	**.00261**	**-.00047**	**.03391**	**-.00323**
Stddev	.00029	.00036	.00083	.00135	.02031	.06552	.00128
%RSD	269.77	155.32	469.03	51.632	4346.1	193.24	39.606
#1	-.00002	.00060	-.00019	.00137	-.00573	.10021	-.00211
#2	-.00010	-.00011	.00113	.00242	.02196	-.03081	-.00295
#3	.00044	.00020	-.00041	.00405	-.01763	.03232	-.00462
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.02721**	**.00109**	**-.00056**	**.23253**	**.00049**	**-.00816**	**-.00015**
Stddev	.04069	.00341	.00026	.02524	.00107	.00624	.00442
%RSD	149.52	313.70	46.647	10.856	219.38	76.495	2897.5
#1	.01942	-.00058	-.00035	.24925	-.00074	-.01380	.00125
#2	-.05550	-.00117	-.00047	.20349	.00124	-.00145	-.00510
#3	-.04556	.00501	-.00085	.24484	.00096	-.00923	.00339
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606089607 Acquired: 6/23/2016 21:20:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00194**	F **-.01015**	**.08416**	**.00022**	**-.00011**	**-.00212**	**-.00416**
Stddev	.00173	.00222	.00063	.00064	.00006	.01168	.00343
%RSD	89.180	21.886	.74271	295.50	49.832	550.35	82.554
#1	-.00101	-.00764	.08452	-.00003	-.00011	-.01340	-.00177
#2	-.00087	-.01186	.08451	-.00027	-.00006	.00993	-.00262
#3	-.00393	-.01095	.08344	.00095	-.00017	-.00289	-.00809
Check ?	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit		90.000					
Low Limit		-.01000					

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00083**	**.00158**	**.28593**
Stddev	.00029	.00007	.07024
%RSD	34.545	4.6737	24.566
#1	-.00055	.00150	.21816
#2	-.00081	.00165	.28123
#3	-.00112	.00159	.35840
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6700.8**	**41736.**	**2133.1**
Stddev	34.6	206.	28.2
%RSD	.51651	.49285	1.3235
#1	6660.9	41521.	2103.9
#2	6723.0	41931.	2160.2
#3	6718.5	41757.	2135.2

Sample Name: L1606082202 Acquired: 6/23/2016 21:24:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00194**	**-.00446**	**.00305**	**.00265**	**.00700**	**.00002**	**.35406**
Stddev	.00134	.00435	.00421	.00519	.00205	.00001	.05184
%RSD	68.787	97.453	137.81	196.12	29.341	37.190	14.641
#1	.00075	-.00690	-.00162	.00667	.00636	.00002	.30481
#2	.00338	.00056	.00423	-.00321	.00930	.00002	.40815
#3	.00169	-.00704	.00655	.00448	.00534	.00003	.34922
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00007**	**.00043**	**-.00077**	**.00285**	**.14848**	**-.04278**	**.00212**
Stddev	.00027	.00044	.00098	.00102	.00792	.02901	.00038
%RSD	394.25	102.19	127.46	35.949	5.3356	67.810	17.947
#1	-.00009	.00000	-.00184	.00239	.15626	-.06821	.00218
#2	.00038	.00041	.00008	.00403	.14874	-.04895	.00171
#3	-.00008	.00089	-.00054	.00214	.14043	-.01118	.00246
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.28223**	**.00480**	**-.00038**	**4.3287**	**-.00014**	**.01612**	**.00046**
Stddev	.10559	.00363	.00060	.1420	.00015	.00579	.00219
%RSD	37.411	75.632	157.86	3.2805	107.54	35.905	474.81
#1	.24611	.00764	-.00099	4.3698	-.00016	.01856	-.00192
#2	.19945	.00071	.00020	4.4456	-.00027	.02028	.00239
#3	.40114	.00605	-.00035	4.1706	.00002	.00951	.00092
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606082202 Acquired: 6/23/2016 21:24:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00001**	**-.00096**	**.60016**	**-.00066**	**.01034**	**-.00345**	**-.00121**
Stddev	.00119	.00324	.00463	.00053	.00056	.00260	.00164
%RSD	9663.9	335.88	.77167	79.904	5.4479	75.303	135.44
#1	-.00094	.00170	.59620	-.00020	.00979	-.00115	-.00170
#2	.00134	-.00002	.59904	-.00124	.01092	-.00293	.00062
#3	-.00037	-.00457	.60525	-.00055	.01031	-.00627	-.00255
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.00034**	**.00082**	F **-.14423**
Stddev	.00011	.00006	.30331
%RSD	33.594	7.1986	210.30
#1	.00036	.00078	-.38581
#2	.00045	.00089	-.24305
#3	.00022	.00079	.19617
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6775.7**	**42604.**	**2060.5**
Stddev	51.6	45.	84.1
%RSD	.76190	.10471	4.0814
#1	6799.5	42589.	2047.0
#2	6811.2	42654.	1984.0
#3	6716.5	42568.	2150.6

Sample Name: L1606082203 Acquired: 6/23/2016 21:28:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00145	.00184	.00074	-.00019	.03852	.00003	1.5262	.00007
Stddev	.00122	.00469	.00175	.00116	.00077	.00012	.0464	.00017
%RSD	84.264	255.24	237.52	615.42	2.0040	388.98	3.0406	258.31
#1	.00264	-.00346	.00229	-.00083	.03875	.00010	1.5782	-.00003
#2	.00020	.00352	.00109	-.00089	.03915	.00011	1.4889	-.00004
#3	.00151	.00546	-.00116	.00115	.03766	-.00011	1.5116	.00026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00037	.00018	.00108	.25045	2.4536	.00344	.67766	.01002
Stddev	.00037	.00129	.00050	.01436	.0243	.00221	.03143	.00186
%RSD	100.70	701.59	46.418	5.7329	.98953	64.075	4.6384	18.607
#1	.00035	.00006	.00058	.23440	2.4813	.00332	.65654	.01217
#2	.00075	.00153	.00159	.26208	2.4433	.00571	.71379	.00880
#3	.00001	-.00104	.00108	.25488	2.4361	.00130	.66266	.00909
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00035	4.5035	-.00023	.01290	-.00147	.00479	-.00584	.55304
Stddev	.00012	.0716	.00036	.00146	.00189	.00145	.00099	.00147
%RSD	35.728	1.5904	157.06	11.351	128.61	30.358	16.944	.26494
#1	-.00047	4.5590	.00019	.01295	-.00040	.00482	-.00623	.55146
#2	-.00035	4.5290	-.00045	.01141	-.00366	.00623	-.00472	.55435
#3	-.00023	4.4227	-.00042	.01433	-.00036	.00332	-.00658	.55332
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606082203 Acquired: 6/23/2016 21:28:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00002	.03922	-.00354	-.00320	-.00104	.00103	-.02734
Stddev	.00093	.00042	.00540	.00176	.00049	.00021	.22820
%RSD	4945.8	1.0630	152.76	54.899	47.016	20.853	834.62
#1	.00001	.03970	-.00944	-.00516	-.00118	.00102	.23169
#2	.00095	.03900	.00116	-.00265	-.00144	.00082	-.19873
#3	-.00091	.03895	-.00233	-.00178	-.00049	.00125	-.11499
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6687.3	41441.	2081.5
Stddev	25.3	113.	22.3
%RSD	.37870	.27264	1.0699
#1	6658.6	41368.	2073.0
#2	6706.4	41384.	2064.8
#3	6696.9	41571.	2106.8

Sample Name: L1606082302 Acquired: 6/23/2016 21:32:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00042	.00297	.00273	.00225	.36313	-.00000	11.836
Stddev	.00174	.00449	.00180	.00282	.00235	.00016	.075
%RSD	415.32	151.16	65.994	125.04	.64751	6744.5	.63384
#1	-.00148	.00121	.00168	.00506	.36408	.00008	11.919
#2	.00080	.00807	.00170	.00226	.36486	-.00018	11.773
#3	.00194	-.00038	.00482	-.00057	.36045	.00010	11.815
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00016	.00291	.00028	.00150	.22899	.13342	.00323
Stddev	.00018	.00030	.00095	.00047	.02084	.11756	.00252
%RSD	113.97	10.305	343.96	31.522	9.1002	88.111	78.086
#1	-.00018	.00302	-.00007	.00192	.21561	.24582	.00054
#2	.00003	.00257	-.00045	.00158	.21836	.01130	.00554
#3	-.00032	.00314	.00135	.00099	.25300	.14315	.00359
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.3954	.04077	-.00002	33.177	.00094	.00875	-.00140
Stddev	.0055	.00391	.00017	.158	.00081	.00388	.00293
%RSD	.05884	9.5814	818.90	.47682	86.099	44.289	209.14
#1	9.3987	.04528	.00014	33.360	.00025	.00895	-.00476
#2	9.3986	.03829	-.00000	33.085	.00074	.00478	.00055
#3	9.3891	.03876	-.00020	33.087	.00184	.01252	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606082302 Acquired: 6/23/2016 21:32:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00211	F -.01027	.87158	-.00061	.28366	-.00294	-.00367
Stddev	.00335	.00460	.01165	.00033	.00032	.00246	.00356
%RSD	158.22	44.796	1.3368	53.749	.11189	83.646	96.934
#1	.00240	-.01556	.85944	-.00095	.28358	-.00225	-.00028
#2	.00531	-.00804	.87263	-.00029	.28401	-.00568	-.00738
#3	-.00136	-.00721	.88268	-.00060	.28339	-.00091	-.00335
Check ?	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit		90.000					
Low Limit		-.01000					

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	-.00036	.00115	.04430
Stddev	.00028	.00013	.31486
%RSD	77.980	11.657	710.78
#1	-.00058	.00122	.37480
#2	-.00005	.00100	.01024
#3	-.00044	.00124	-.25215
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6666.9	40963.	2080.0
Stddev	54.9	248.	35.1
%RSD	.82321	.60425	1.6856
#1	6605.2	41025.	2039.5
#2	6685.2	40690.	2100.5
#3	6710.3	41173.	2100.0

Sample Name: CCV Acquired: 6/23/2016 21:36:11 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.37886	9.4008	.36977	.46580	.94543	.04748	9.4956
Stddev	.00192	.0240	.00492	.00165	.00279	.00009	.0298
%RSD	.50559	.25544	1.3304	.35366	.29493	.18359	.31411
#1	.37848	9.4279	.37519	.46422	.94745	.04749	9.4832
#2	.38093	9.3822	.36851	.46567	.94658	.04756	9.4739
#3	.37716	9.3923	.36560	.46751	.94225	.04739	9.5296
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04722	.18900	.47404	.47679	3.8147	47.440	.96725
Stddev	.00010	.00048	.00128	.00128	.0064	.112	.00455
%RSD	.21398	.25239	.26979	.26927	.16823	.23504	.47080
#1	.04722	.18886	.47544	.47536	3.8073	47.493	.96637
#2	.04711	.18861	.47292	.47785	3.8188	47.516	.97218
#3	.04731	.18953	.47377	.47715	3.8179	47.312	.96320
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.5275	.47798	.95150	47.323	.47305	9.3519	.47463
Stddev	.0772	.00629	.00356	.158	.00089	.0108	.00368
%RSD	.81075	1.3163	.37366	.33377	.18727	.11544	.77586
#1	9.5741	.48274	.95557	47.363	.47379	9.3402	.47038
#2	9.4383	.47085	.94902	47.457	.47331	9.3538	.47650
#3	9.5701	.48036	.94991	47.149	.47207	9.3616	.47699
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 21:36:11 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1248**	**.37983**	**4.7469**	**.94195**	**.94936**	**.95013**	**.47646**
Stddev	.0071	.00577	.0107	.00244	.00345	.00861	.00052
%RSD	.63160	1.5193	.22491	.25903	.36313	.90637	.10952
#1	1.1176	.38628	4.7372	.93943	.94930	.94189	.47702
#2	1.1251	.37515	4.7451	.94210	.95284	.94943	.47599
#3	1.1318	.37806	4.7583	.94431	.94595	.95907	.47637
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.94937**	**.94698**	F **.53390**
Stddev	.00172	.00064	.57416
%RSD	.18154	.06770	107.54
#1	.95103	.94624	-.12311
#2	.94951	.94740	.93934
#3	.94759	.94730	.78547
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6847.6**	**42330.**	**2179.9**
Stddev	26.3	36.	23.0
%RSD	.38348	.08457	1.0562
#1	6817.2	42321.	2206.4
#2	6862.7	42299.	2165.2
#3	6862.8	42369.	2168.0

Sample Name: CCB Acquired: 6/23/2016 21:39:49 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00034**	**-.00840**	**.00142**	**.00310**	**-.00097**	**-.00004**	**.04289**
Stddev	.00215	.00740	.00154	.00403	.00114	.00009	.02517
%RSD	634.22	88.117	107.88	130.11	117.13	211.10	58.687
#1	.00114	-.00019	.00313	.00134	-.00178	-.00003	.05755
#2	.00065	-.01045	.00097	.00025	-.00148	.00004	.01383
#3	-.00280	-.01457	.00016	.00772	.00033	-.00014	.05730
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00014**	**.00024**	**-.00021**	**.00099**	**.00681**	**.02724**	**.00474**
Stddev	.00017	.00030	.00131	.00197	.01403	.13786	.00231
%RSD	120.69	124.53	625.16	199.14	206.17	506.05	48.675
#1	-.00003	.00049	-.00143	.00036	-.00450	-.09799	.00738
#2	.00014	.00033	-.00037	.00320	.02251	.00476	.00369
#3	.00030	-.00009	.00117	-.00059	.00241	.17496	.00314
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.01627**	**.00349**	**.00316**	**.05656**	**.00129**	**.00259**	**.00220**
Stddev	.06560	.00061	.00054	.03460	.00090	.00126	.00512
%RSD	403.12	17.500	17.006	61.166	69.213	48.482	232.87
#1	.04308	.00418	.00337	.06648	.00145	.00133	.00302
#2	-.00519	.00302	.00356	.01809	.00033	.00384	-.00328
#3	-.08670	.00326	.00255	.08512	.00210	.00261	.00686
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 21:39:49 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00210**	**.00060**	**-.00131**	**-.00025**	**.00045**	**.00101**	**-.00537**
Stddev	.00282	.00202	.00149	.00144	.00024	.00715	.00270
%RSD	134.10	334.37	113.55	585.99	53.933	705.65	50.282
#1	.00467	-.00103	-.00073	-.00138	.00029	-.00721	-.00765
#2	-.00091	.00286	-.00300	-.00074	.00033	.00583	-.00608
#3	.00255	-.00002	-.00020	.00138	.00073	.00442	-.00239
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00036**	**.00019**	F **.41711**
Stddev	.00059	.00027	.26410
%RSD	164.24	139.43	63.317
#1	-.00011	-.00011	.71541
#2	.00007	.00039	.21306
#3	-.00103	.00030	.32286
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6996.3**	**43082.**	**2193.8**
Stddev	20.8	164.	36.3
%RSD	.29721	.37998	1.6551
#1	6972.3	42895.	2225.5
#2	7007.7	43194.	2201.8
#3	7008.9	43159.	2154.2

Sample Name: L1606082303 Acquired: 6/23/2016 21:43:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00032	.00528	-.00158	.00205	.19615	.00000	9.5412	-.00005
Stddev	.00172	.00389	.00374	.00056	.00167	.00004	.0635	.00015
%RSD	531.26	73.680	236.25	27.382	.84980	2856.4	.66503	295.12
#1	.00039	.00192	-.00320	.00177	.19699	.00004	9.6087	.00011
#2	.00201	.00955	-.00424	.00168	.19423	-.00003	9.4828	-.00019
#3	-.00143	.00438	.00269	.00269	.19724	-.00001	9.5322	-.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	-.00003	.00137	.01816	.07360	-.00201	5.7035	.11460
Stddev	.00028	.00217	.00077	.02760	.04149	.00162	.1926	.00292
%RSD	57.802	7719.6	56.083	151.98	56.373	80.472	3.3772	2.5499
#1	.00048	-.00239	.00048	.04879	.11308	-.00147	5.9188	.11188
#2	.00020	.00186	.00178	-.00477	.07737	-.00383	5.5474	.11423
#3	.00076	.00045	.00184	.01046	.03036	-.00073	5.6443	.11769
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00025	30.457	.00028	.01105	.00022	.00112	-.00912	.60825
Stddev	.00054	.348	.00034	.00162	.00227	.00456	.00372	.00164
%RSD	210.59	1.1437	122.90	14.679	1048.7	408.61	40.764	.26908
#1	-.00033	30.637	-.00010	.01036	-.00046	-.00379	-.00828	.60645
#2	.00038	30.056	.00057	.01291	.00274	.00192	-.00590	.60965
#3	.00071	30.679	.00036	.00990	-.00164	.00522	-.01319	.60865
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606082303 Acquired: 6/23/2016 21:43:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00070**	**.18364**	**-.00297**	**-.00161**	**.00010**	**.00105**	**.18385**
Stddev	.00086	.00144	.00537	.00137	.00034	.00025	.28076
%RSD	122.13	.78353	180.76	84.915	342.41	23.814	152.71
#1	-.00168	.18402	-.00642	-.00247	-.00015	.00082	-.00761
#2	-.00010	.18205	.00322	-.00003	-.00005	.00103	.50616
#3	-.00032	.18485	-.00570	-.00232	.00050	.00132	.05301
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6737.6**	**41321.**	**2154.5**
Stddev	10.1	398.	10.5
%RSD	.15056	.96263	.48515
#1	6747.5	41745.	2162.1
#2	6727.2	41262.	2142.5
#3	6738.0	40956.	2158.7

Sample Name: L1606082304 Acquired: 6/23/2016 21:47:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00107	.00763	.00205	.00128	.06210	-.00005	2.2162	-.00002
Stddev	.00070	.00357	.00361	.00281	.00225	.00003	.0545	.00005
%RSD	65.790	46.743	175.75	220.17	3.6308	56.826	2.4590	342.73
#1	.00064	.01174	.00606	-.00197	.06015	-.00004	2.1625	-.00007
#2	.00188	.00577	-.00095	.00288	.06159	-.00009	2.2715	.00004
#3	.00068	.00537	.00106	.00292	.06457	-.00003	2.2145	-.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	-.00139	.00241	4.0573	.28730	.00089	1.4304	.07372
Stddev	.00015	.00065	.00132	.0516	.13779	.00299	.0553	.00403
%RSD	30.727	46.428	54.848	1.2708	47.960	334.26	3.8623	5.4701
#1	.00066	-.00136	.00091	4.0586	.13895	.00365	1.4701	.07443
#2	.00040	-.00206	.00293	4.0051	.41127	.00130	1.4537	.07736
#3	.00040	-.00076	.00340	4.1081	.31169	-.00228	1.3673	.06938
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00002	6.7673	.00003	.01064	-.00183	.00117	-.00224	.82027
Stddev	.00026	.0658	.00066	.00541	.00174	.00605	.00594	.00641
%RSD	1414.1	.97251	2005.0	50.852	95.153	515.39	265.01	.78184
#1	-.00017	6.7130	.00068	.01686	-.00367	.00122	-.00440	.81299
#2	-.00008	6.7484	.00006	.00808	-.00161	-.00490	-.00681	.82274
#3	.00031	6.8405	-.00064	.00698	-.00021	.00720	.00448	.82508
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606082304 Acquired: 6/23/2016 21:47:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00062**	**.05041**	**.00372**	**-.00162**	**-.00102**	**.00139**	**.02709**
Stddev	.00054	.00053	.00900	.00223	.00128	.00033	.20037
%RSD	86.087	1.0604	241.81	137.59	125.65	23.588	739.53
#1	-.00016	.04984	-.00362	-.00387	-.00076	.00168	-.08640
#2	-.00121	.05049	.01377	-.00158	-.00240	.00104	.25845
#3	-.00050	.05090	.00102	.00059	.00011	.00145	-.09076
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6801.0**	**42034.**	**2119.8**
Stddev	49.8	142.	31.5
%RSD	.73210	.33806	1.4863
#1	6763.5	42178.	2123.4
#2	6782.1	42032.	2149.4
#3	6857.5	41894.	2086.7

Sample Name: L1606088302 Acquired: 6/23/2016 21:51:41 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00081	.03900	.00448	.00430	.01505	-.00001	2.8996	-.00001
Stddev	.00124	.00075	.00243	.00233	.00060	.00004	.0350	.00023
%RSD	153.65	1.9305	54.189	54.088	3.9770	332.40	1.2074	2343.8
#1	-.00027	.03833	.00184	.00580	.01560	.00003	2.9325	.00014
#2	.00216	.03982	.00662	.00548	.01441	-.00003	2.8628	.00010
#3	.00053	.03886	.00496	.00162	.01514	-.00004	2.9035	-.00027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00051	-.00093	.00225	.26018	.53798	-.00133	1.2540	.13015
Stddev	.00005	.00064	.00096	.01052	.04447	.00237	.1006	.00182
%RSD	10.629	68.468	42.462	4.0434	8.2670	177.47	8.0229	1.3964
#1	.00052	-.00155	.00287	.26853	.50853	-.00040	1.3655	.12853
#2	.00056	-.00028	.00273	.26365	.51627	-.00403	1.2265	.12979
#3	.00045	-.00096	.00115	.24836	.58914	.00042	1.1700	.13212
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00023	2.6639	.00090	.01608	.00052	.00237	-.00249	.36519
Stddev	.00017	.0233	.00047	.00836	.00171	.00049	.00862	.00186
%RSD	72.830	.87410	52.415	51.991	330.96	20.707	346.32	.51020
#1	.00029	2.6639	.00099	.02573	.00209	.00189	.00736	.36326
#2	.00036	2.6406	.00039	.01116	.00078	.00287	-.00614	.36532
#3	.00004	2.6871	.00132	.01135	-.00131	.00234	-.00868	.36698
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606088302 Acquired: 6/23/2016 21:51:41 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00006**	**.01730**	**.00326**	**-.00074**	**-.00020**	**.00699**	**.35536**
Stddev	.00047	.00025	.00678	.00023	.00043	.00024	.51910
%RSD	826.36	1.4568	208.35	30.989	219.10	3.4850	146.08
#1	.00036	.01713	.00871	-.00090	-.00020	.00693	-.23473
#2	-.00056	.01718	-.00434	-.00084	-.00062	.00678	.74156
#3	.00004	.01759	.00541	-.00048	.00024	.00726	.55925
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**7019.0**	**43469.**	**2272.3**
Stddev	25.4	199.	24.4
%RSD	.36132	.45762	1.0740
#1	7020.2	43542.	2267.1
#2	6993.1	43621.	2298.9
#3	7043.7	43244.	2250.9

Sample Name: L1606089601 Acquired: 6/23/2016 21:55:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00420**	**.00523**	**.00334**	**.00237**	**.00219**	**.00008**	**.19980**
Stddev	.00278	.00696	.00261	.00237	.00063	.00004	.02591
%RSD	66.297	133.09	78.109	99.957	28.753	51.900	12.967
#1	.00741	.00603	.00348	.00033	.00147	.00009	.18026
#2	.00260	.01174	.00066	.00182	.00265	.00012	.22918
#3	.00258	-.00210	.00587	.00496	.00244	.00003	.18994
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00015**	**.00302**	**.00023**	**.00052**	**2.0027**	**-.01208**	**-.00040**
Stddev	.00012	.00024	.00052	.00149	.0311	.05532	.00593
%RSD	79.007	7.8027	222.58	284.84	1.5506	458.05	1480.9
#1	-.00011	.00287	.00040	.00177	1.9805	.05152	-.00717
#2	-.00005	.00329	-.00035	-.00112	1.9893	-.03870	.00382
#3	-.00028	.00290	.00064	.00091	2.0382	-.04905	.00215
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.24211**	**.28328**	**-.00023**	**.95184**	**.00021**	**.00495**	**-.00083**
Stddev	.13780	.01030	.00015	.04331	.00103	.00358	.00274
%RSD	56.917	3.6352	65.757	4.5498	490.79	72.184	329.16
#1	.26334	.27250	-.00010	.90187	.00062	.00759	.00169
#2	.09492	.28431	-.00019	.97855	.00097	.00088	-.00043
#3	.36806	.29302	-.00040	.97509	-.00096	.00638	-.00376
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606089601 Acquired: 6/23/2016 21:55:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00174**	**-.00552**	**.20894**	**-.00089**	**.00183**	**.00041**	**-.00355**
Stddev	.00455	.00357	.00152	.00075	.00028	.00397	.00196
%RSD	261.01	64.670	.72600	84.586	15.398	961.73	55.133
#1	-.00696	-.00337	.20831	-.00159	.00215	.00126	-.00263
#2	.00141	-.00964	.20784	-.00009	.00174	.00389	-.00222
#3	.00032	-.00355	.21067	-.00099	.00161	-.00391	-.00580
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00023**	**.00107**	F **-.16350**
Stddev	.00118	.00009	.38488
%RSD	512.00	8.7488	235.40
#1	-.00072	.00105	.00228
#2	-.00109	.00099	.11071
#3	.00111	.00117	-.60348
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6828.9**	**41943.**	**2135.6**
Stddev	63.3	286.	20.3
%RSD	.92751	.68281	.94968
#1	6763.6	42271.	2138.9
#2	6833.1	41741.	2154.0
#3	6890.1	41817.	2113.9

Sample Name: L1606089602 Acquired: 6/23/2016 21:59:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00116	.00173	.00317	.00184	-.00011	.00003	.00624	-.00003
Stddev	.00285	.00778	.00249	.00329	.00125	.00006	.05063	.00023
%RSD	245.09	449.18	78.692	179.21	1183.1	208.66	811.35	713.72
#1	-.00055	.01019	.00053	.00389	-.00026	-.00003	-.04528	.00022
#2	.00445	-.00512	.00349	.00357	-.00127	.00009	.05593	-.00023
#3	-.00041	.00012	.00549	-.00196	.00122	.00003	.00807	-.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00024	.00046	.00165	-.00531	.07219	-.00031	.05667	.00209
Stddev	.00028	.00102	.00160	.03109	.09458	.00388	.05878	.00182
%RSD	116.13	222.95	97.271	585.63	131.01	1232.5	103.73	87.087
#1	.00001	.00001	.00226	.02472	.15298	-.00298	.09347	.00378
#2	.00016	-.00027	-.00017	-.00328	.09542	.00413	.08765	.00016
#3	.00055	.00163	.00285	-.03736	-.03184	-.00209	-.01113	.00234
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00056	.39833	-.00077	.00469	.00112	.00031	-.00922	.46168
Stddev	.00006	.02012	.00058	.00167	.00277	.00100	.00659	.00725
%RSD	11.375	5.0505	75.062	35.563	245.83	319.68	71.432	1.5707
#1	-.00061	.37822	-.00128	.00414	.00250	-.00075	-.01634	.45603
#2	-.00049	.39832	-.00088	.00656	-.00206	.00124	-.00799	.45917
#3	-.00057	.41846	-.00014	.00336	.00293	.00045	-.00334	.46986
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089602 Acquired: 6/23/2016 21:59:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00058**	**.00007**	**-.00373**	**-.00407**	**-.00032**	**.00100**	**-.02035**
Stddev	.00028	.00019	.00957	.00088	.00078	.00036	.45457
%RSD	48.712	259.28	256.71	21.746	243.98	36.577	2234.1
#1	-.00053	.00018	.00727	-.00462	.00006	.00142	-.06540
#2	-.00033	.00018	-.00832	-.00454	.00020	.00080	.45507
#3	-.00089	-.00014	-.01012	-.00305	-.00121	.00077	-.45071
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6764.1**	**42193.**	**2152.3**
Stddev	51.5	150.	58.2
%RSD	.76176	.35597	2.7055
#1	6724.7	42019.	2097.0
#2	6745.2	42282.	2213.0
#3	6822.4	42277.	2146.9

Sample Name: L1606089603 Acquired: 6/23/2016 22:03:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00264**	**.00052**	**.00129**	**.00105**	**.00283**	**.00007**	**.13087**	**-.00015**
Stddev	.00174	.01588	.00262	.00317	.00277	.00004	.02295	.00021
%RSD	65.852	3062.5	203.49	301.32	97.964	60.620	17.533	135.33
#1	.00093	.01257	-.00138	.00463	.00302	.00005	.14125	.00004
#2	.00440	.00646	.00387	-.00005	.00549	.00004	.10457	-.00037
#3	.00259	-.01748	.00137	-.00142	-.00004	.00011	.14679	-.00013
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00048**	**.00024**	**.00128**	**.12772**	**.20105**	**-.00470**	**.12553**	**.00043**
Stddev	.00031	.00150	.00216	.02497	.11408	.00639	.06203	.00287
%RSD	63.789	616.72	168.54	19.548	56.742	136.10	49.418	664.80
#1	.00045	.00194	-.00101	.11916	.09190	.00173	.17340	-.00027
#2	.00019	-.00030	.00156	.10815	.19176	-.00476	.14774	-.00202
#3	.00080	-.00091	.00330	.15584	.31949	-.01106	.05545	.00359
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00003**	**.39363**	**-.00055**	**.01390**	**-.00151**	**.00183**	**-.00632**	**.24727**
Stddev	.00045	.04518	.00102	.00707	.00255	.00239	.00219	.00127
%RSD	1327.8	11.477	186.18	50.889	169.36	131.03	34.571	.51246
#1	-.00019	.43317	-.00107	.01491	-.00330	.00328	-.00397	.24716
#2	.00047	.34439	.00063	.02042	.00142	-.00094	-.00828	.24858
#3	-.00038	.40334	-.00121	.00638	-.00264	.00314	-.00672	.24606
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089603 Acquired: 6/23/2016 22:03:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00047**	**.00125**	**-.01273**	**-.00400**	**-.00003**	**.00095**	**.10691**
Stddev	.00077	.00042	.00925	.00266	.00030	.00011	.21605
%RSD	161.90	33.664	72.615	66.501	1119.1	11.082	202.08
#1	-.00019	.00148	-.00295	-.00574	-.00008	.00101	-.14101
#2	-.00134	.00076	-.01391	-.00094	.00030	.00102	.20680
#3	.00011	.00150	-.02133	-.00533	-.00030	.00083	.25493
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6778.0**	**41534.**	**2155.4**
Stddev	33.3	392.	34.0
%RSD	.49153	.94350	1.5766
#1	6804.2	41394.	2116.9
#2	6740.5	41232.	2181.3
#3	6789.2	41977.	2167.8

Sample Name: L1606089604 Acquired: 6/23/2016 22:07:36 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00159	.01026	.00044	.00222	.00123	.00002	.21290	.00018
Stddev	.00043	.01032	.00191	.00166	.00091	.00006	.03422	.00024
%RSD	27.269	100.60	437.55	74.707	74.454	271.36	16.075	135.58
#1	.00148	.01567	.00246	.00277	.00044	.00008	.21883	-.00010
#2	.00207	.01676	-.00133	.00035	.00223	.00002	.24377	.00035
#3	.00122	-.00164	.00018	.00353	.00102	-.00004	.17610	.00029
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00372	-.00098	.00204	1.9871	.04803	-.00029	.27775	.28079
Stddev	.00019	.00146	.00258	.0500	.09028	.00333	.05841	.00512
%RSD	4.9815	148.68	126.42	2.5186	187.98	1134.6	21.028	1.8227
#1	.00391	-.00254	-.00094	1.9637	-.03733	-.00218	.34159	.28507
#2	.00373	.00037	.00345	1.9529	.03888	.00355	.22701	.27512
#3	.00354	-.00079	.00362	2.0445	.14254	-.00224	.26465	.28217
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00026	.88820	.00047	.00405	-.00241	-.00066	-.00911	.20760
Stddev	.00009	.05397	.00049	.00324	.00083	.00328	.00427	.00263
%RSD	33.170	6.0764	104.78	80.127	34.563	498.25	46.914	1.2688
#1	-.00017	.85983	.00038	.00127	-.00145	.00099	-.00982	.20605
#2	-.00027	.85432	.00100	.00326	-.00287	.00147	-.00453	.20610
#3	-.00035	.95043	.00003	.00761	-.00292	-.00443	-.01299	.21064
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089604 Acquired: 6/23/2016 22:07:36 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**.00180**	**.00146**	**-.00008**	**.00002**	**.00083**	**.01606**
Stddev	.00033	.00028	.00448	.00273	.00107	.00008	.15859
%RSD	645.35	15.485	305.81	3306.5	5321.6	9.5646	987.51
#1	-.00043	.00167	.00571	-.00323	-.00007	.00090	-.04948
#2	.00011	.00212	.00190	.00129	-.00100	.00085	-.09926
#3	.00017	.00161	-.00322	.00169	.00113	.00074	.19692
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6852.5**	**41879.**	**2133.3**
Stddev	55.4	422.	31.4
%RSD	.80848	1.0073	1.4729
#1	6793.1	41503.	2112.7
#2	6861.8	42335.	2169.5
#3	6902.7	41799.	2117.7

Sample Name: L1606089605 Acquired: 6/23/2016 22:11:34 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00325**	**-.00813**	**.00321**	**.00103**	**.00232**	**.00007**	**.13232**
Stddev	.00086	.00834	.00166	.00270	.00015	.00005	.04505
%RSD	26.517	102.62	51.687	262.17	6.5199	65.910	34.048
#1	.00357	-.00015	.00448	.00315	.00234	.00011	.13419
#2	.00227	-.00744	.00133	-.00201	.00246	.00002	.08636
#3	.00390	-.01679	.00380	.00194	.00216	.00008	.17641
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00022**	**.00019**	**-.00021**	**-.00051**	**.02399**	**.13993**	**-.00272**
Stddev	.00013	.00048	.00060	.00074	.02883	.04909	.00167
%RSD	59.092	248.65	278.22	144.43	120.17	35.082	61.414
#1	-.00034	.00000	-.00032	-.00045	-.00884	.19045	-.00464
#2	-.00023	.00074	.00043	-.00128	.04519	.13693	-.00190
#3	-.00008	-.00016	-.00075	.00019	.03563	.09241	-.00161
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.13116**	**-.00087**	**-.00017**	**.35584**	**.00083**	**.00721**	**.00009**
Stddev	.12028	.00028	.00012	.03175	.00083	.00389	.00231
%RSD	91.707	32.079	72.680	8.9233	100.69	53.875	2586.3
#1	.15231	-.00055	-.00032	.37833	.00083	.00320	-.00213
#2	.23946	-.00098	-.00010	.31952	-.00001	.01096	.00248
#3	.00170	-.00108	-.00010	.36967	.00166	.00747	-.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: L1606089605 Acquired: 6/23/2016 22:11:34 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00575**	**.00189**	**.24045**	**-.00028**	**.00101**	**-.00455**	**-.00361**
Stddev	.00192	.00588	.00568	.00092	.00006	.01057	.00467
%RSD	33.402	311.30	2.3620	328.38	6.1198	232.48	129.52
#1	.00788	.00697	.23736	-.00119	.00106	-.01626	-.00722
#2	.00521	.00325	.23698	.00066	.00103	-.00167	-.00528
#3	.00416	-.00455	.24700	-.00032	.00094	.00429	.00167
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00106**	**.00104**	F **-.05759**
Stddev	.00074	.00009	.21862
%RSD	69.542	8.5436	379.61
#1	-.00181	.00101	.17219
#2	-.00033	.00097	-.26300
#3	-.00105	.00114	-.08197
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6837.2**	**42004.**	**2152.4**
Stddev	16.0	173.	9.0
%RSD	.23440	.41281	.41704
#1	6853.2	42014.	2142.1
#2	6821.1	41826.	2156.7
#3	6837.2	42172.	2158.4

Sample Name: L1606089607 Acquired: 6/23/2016 22:15:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00124**	**-.00686**	**.00409**	**.00130**	**-.00066**	**-.00005**	**.03206**	**-.00002**
Stddev	.00132	.01058	.00305	.00213	.00072	.00007	.01681	.00018
%RSD	105.94	154.08	74.558	163.59	109.05	125.85	52.443	856.76
#1	-.00264	-.01401	.00622	.00259	-.00111	-.00005	.04548	-.00001
#2	-.00105	-.01187	.00546	.00246	-.00105	.00001	.01320	.00015
#3	-.00003	.00529	.00060	-.00115	.00017	-.00012	.03751	-.00021
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00031**	**.00074**	**.00061**	**-.00910**	**-.00529**	**-.00407**	**.02444**	**.00176**
Stddev	.00024	.00180	.00036	.03159	.06755	.00591	.05716	.00378
%RSD	78.629	244.31	60.097	347.19	1277.4	145.34	233.90	215.59
#1	.00037	.00281	.00094	.01641	.06833	-.00447	-.04095	.00014
#2	.00051	-.00035	.00022	-.04443	-.01978	-.00977	.04932	-.00095
#3	.00004	-.00025	.00066	.00073	-.06441	.00203	.06494	.00608
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00073**	**.00595**	**.00025**	**.00693**	**-.00226**	**-.00313**	**-.00892**	**-.00746**
Stddev	.00042	.03243	.00102	.00571	.00047	.00121	.00509	.00076
%RSD	57.397	545.12	410.30	82.361	20.886	38.630	57.024	10.142
#1	-.00072	.04302	.00024	.00636	-.00177	-.00362	-.00954	-.00789
#2	-.00116	-.01720	-.00077	.01291	-.00271	-.00175	-.01368	-.00792
#3	-.00032	-.00797	.00128	.00153	-.00229	-.00401	-.00356	-.00659
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606089607 Acquired: 6/23/2016 22:15:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00050**	**.00006**	**-.00267**	**-.00165**	**-.00070**	**.00059**	**.18991**
Stddev	.00084	.00062	.00845	.00365	.00135	.00010	.03400
%RSD	167.93	1027.2	316.03	221.52	192.73	16.308	17.901
#1	-.00037	-.00011	.00072	-.00580	.00084	.00058	.15066
#2	-.00140	.00075	-.01229	-.00022	-.00169	.00050	.20942
#3	.00027	-.00046	.00355	.00107	-.00126	.00069	.20967
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6854.9**	**42369.**	**2139.0**
Stddev	37.2	282.	6.9
%RSD	.54331	.66600	.32457
#1	6813.1	42056.	2146.4
#2	6867.1	42446.	2132.6
#3	6884.6	42604.	2138.0

Sample Name: CCV Acquired: 6/23/2016 22:19:31 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.37294	9.2890	.37002	.46325	.93343	.04671	9.3219
Stddev	.00058	.0248	.00029	.00298	.00858	.00002	.0885
%RSD	.15646	.26680	.07810	.64298	.91954	.04347	.94918
#1	.37306	9.3173	.36998	.46028	.92385	.04673	9.2510
#2	.37230	9.2715	.36976	.46623	.94043	.04670	9.2937
#3	.37345	9.2781	.37033	.46323	.93599	.04670	9.4210
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04629	.18580	.46699	.46774	3.7375	46.911	.95278
Stddev	.00023	.00060	.00144	.00134	.0018	.467	.00684
%RSD	.48818	.32069	.30929	.28620	.04690	.99443	.71816
#1	.04654	.18545	.46662	.46619	3.7368	46.377	.94493
#2	.04624	.18649	.46859	.46843	3.7395	47.239	.95591
#3	.04609	.18546	.46577	.46859	3.7362	47.116	.95750
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.2808	.46805	.93442	46.553	.46467	9.1850	.46445
Stddev	.1781	.00201	.00336	.432	.00229	.0141	.00099
%RSD	1.9192	.42908	.36011	.92728	.49257	.15361	.21367
#1	9.0878	.46660	.93610	46.055	.46543	9.1763	.46458
#2	9.4388	.47034	.93661	46.821	.46649	9.2013	.46537
#3	9.3158	.46721	.93054	46.783	.46210	9.1774	.46340
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 22:19:31 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1058**	**.37399**	**4.6590**	**.92793**	**.93602**	**.93884**	**.47136**
Stddev	.0068	.00623	.0143	.00333	.00809	.01829	.00375
%RSD	.61331	1.6668	.30694	.35899	.86439	1.9481	.79529
#1	1.1040	.37143	4.6586	.92611	.92682	.91846	.46721
#2	1.1133	.38110	4.6734	.93177	.94202	.94427	.47237
#3	1.1001	.36945	4.6448	.92590	.93921	.95381	.47450
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.93328**	**.93064**	F **.30272**
Stddev	.00145	.00126	.34844
%RSD	.15486	.13568	115.11
#1	.93249	.92922	.70352
#2	.93240	.93165	.07184
#3	.93495	.93106	.13279
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6994.5**	**42945.**	**2201.9**
Stddev	26.0	472.	17.5
%RSD	.37198	1.0980	.79423
#1	7017.9	42835.	2192.4
#2	6966.5	42539.	2222.1
#3	6999.0	43462.	2191.2

Sample Name: CCB Acquired: 6/23/2016 22:23:10 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00075**	**-.01234**	**.00307**	**.00387**	**-.00152**	**.00003**	**.02196**
Stddev	.00074	.00351	.00195	.00329	.00037	.00010	.03751
%RSD	98.761	28.475	63.502	85.024	24.368	329.31	170.81
#1	.00146	-.01497	.00353	.00757	-.00195	-.00000	.04383
#2	.00082	-.00835	.00475	.00124	-.00126	.00014	.04339
#3	-.00002	-.01370	.00093	.00281	-.00136	-.00005	-.02135
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00007**	**.00029**	**-.00074**	**.00053**	**-.01044**	**.06034**	**-.00158**
Stddev	.00023	.00031	.00153	.00009	.01838	.08449	.00205
%RSD	314.24	106.88	205.57	16.843	176.11	140.03	130.11
#1	-.00010	-.00004	-.00241	.00043	-.01522	.02142	-.00392
#2	-.00029	.00057	-.00042	.00060	.00986	.15728	-.00009
#3	.00017	.00033	.00060	.00054	-.02595	.00232	-.00072
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.02604**	**-.00075**	**.00305**	**.06420**	**.00000**	**.00651**	**-.00362**
Stddev	.05654	.00328	.00069	.00709	.00080	.00504	.00339
%RSD	217.15	436.24	22.462	11.042	53218.	77.432	93.854
#1	.03584	-.00047	.00226	.06569	.00068	.00969	-.00632
#2	-.07502	.00238	.00342	.07043	-.00088	.00913	-.00473
#3	-.03894	-.00416	.00347	.05649	.00021	.00070	.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 22:23:10 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00325	-.00859	-.00081	-.00049	.00016	.00095	-.00269
Stddev	.00288	.00100	.00106	.00098	.00044	.00395	.00131
%RSD	88.832	11.667	130.17	199.49	265.70	416.04	48.575
#1	.00425	-.00769	-.00116	-.00120	.00053	-.00220	-.00394
#2	-.00001	-.00968	.00037	.00062	.00028	-.00034	-.00133
#3	.00549	-.00841	-.00166	-.00089	-.00032	.00538	-.00281
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	-.00046	.00011	F .20521
Stddev	.00084	.00022	.40407
%RSD	184.00	201.71	196.91
#1	.00051	-.00006	.57360
#2	-.00094	.00036	-.22696
#3	-.00095	.00003	.26899
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6996.7	43446.	2222.8
Stddev	42.9	110.	21.9
%RSD	.61271	.25379	.98413
#1	6948.6	43485.	2197.6
#2	7030.8	43322.	2237.5
#3	7010.7	43532.	2233.2

Sample Name: LLCCV Acquired: 6/23/2016 22:27:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00907	.13685	.00722	.07327	.00636	.00136	.38737	.00083
Stddev	.00127	.00548	.00269	.00354	.00121	.00002	.02958	.00013
%RSD	14.057	4.0063	37.324	4.8356	18.957	1.6251	7.6350	15.322
#1	.00768	.13064	.00454	.06952	.00698	.00135	.41482	.00073
#2	.01019	.13890	.00718	.07656	.00497	.00135	.39124	.00078
#3	.00934	.14102	.00993	.07374	.00713	.00139	.35605	.00097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00356	.00351	.00643	.06036	.74512	.07304	.35518	.00670
Stddev	.00034	.00083	.00051	.03616	.18336	.00520	.11515	.00028
%RSD	9.6921	23.706	7.9731	59.906	24.608	7.1243	32.420	4.2320
#1	.00383	.00446	.00589	.02265	.76354	.06708	.45336	.00638
#2	.00367	.00316	.00691	.06369	.91857	.07668	.38374	.00683
#3	.00317	.00291	.00648	.09475	.55324	.07536	.22844	.00690
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00743	.43578	.01531	.74167	.00827	.07927	.00569	.77935
Stddev	.00037	.04956	.00080	.00526	.00366	.00493	.00102	.00095
%RSD	5.0056	11.373	5.2515	.70948	44.188	6.2194	17.916	.12251
#1	.00704	.38130	.01503	.74257	.01154	.08395	.00451	.78008
#2	.00779	.47820	.01469	.73602	.00432	.07413	.00628	.77827
#3	.00745	.44783	.01622	.74642	.00896	.07973	.00628	.77970
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: LLCCV Acquired: 6/23/2016 22:27:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.37348	.03865	.02465	.14094	.00676	.01578	19.333
Stddev	.00058	.00060	.00437	.00098	.00113	.00019	.577
%RSD	.15634	1.5628	17.713	.69785	16.700	1.2102	2.9862
#1	.37294	.03795	.02415	.14073	.00773	.01557	19.532
#2	.37410	.03901	.02924	.14008	.00552	.01583	19.783
#3	.37339	.03899	.02055	.14202	.00705	.01594	18.682
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6996.1	43552.	2208.7
Stddev	29.5	407.	29.7
%RSD	.42198	.93392	1.3436
#1	6976.9	43985.	2211.3
#2	7030.1	43494.	2236.9
#3	6981.2	43178.	2177.8

Sample Name: LLCCV Acquired: 6/23/2016 22:31:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02091	.36245	.02162	.18354	.01867	.00356	.98086
Stddev	.00237	.00484	.00558	.00414	.00095	.00018	.04469
%RSD	11.351	1.3361	25.792	2.2543	5.0969	5.1755	4.5561
#1	.01879	.36503	.01522	.18805	.01837	.00338	.94866
#2	.02045	.35687	.02541	.17992	.01791	.00356	.96204
#3	.02347	.36547	.02424	.18267	.01974	.00374	1.0319
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00183	.00959	.00895	.01155	.18702	1.8760	.19064
Stddev	.00040	.00066	.00025	.00083	.02563	.0398	.00709
%RSD	21.803	6.8545	2.7509	7.2203	13.702	2.1189	3.7192
#1	.00229	.01035	.00908	.01200	.20673	1.9212	.19761
#2	.00159	.00922	.00866	.01059	.15805	1.8607	.18344
#3	.00161	.00920	.00910	.01205	.19628	1.8463	.19086
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.90135	.02189	.01808	1.0102	.03771	1.7925	.01852
Stddev	.03529	.00047	.00019	.0359	.00135	.0154	.00112
%RSD	3.9156	2.1507	1.0770	3.5580	3.5742	.86119	6.0288
#1	.87617	.02240	.01801	.97340	.03783	1.7983	.01932
#2	.88620	.02181	.01830	1.0119	.03900	1.8043	.01725
#3	.94169	.02146	.01794	1.0452	.03631	1.7750	.01900
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: LLCCV Acquired: 6/23/2016 22:31:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19437**	**.03462**	**1.8270**	**.92649**	**.09469**	**.05843**	**.34726**
Stddev	.00460	.00283	.0121	.00858	.00060	.00582	.00417
%RSD	2.3690	8.1675	.66332	.92568	.63058	9.9521	1.1997
#1	.19959	.03615	1.8304	.92739	.09439	.06509	.34780
#2	.19266	.03636	1.8370	.93459	.09431	.05586	.35113
#3	.19087	.03136	1.8135	.91750	.09538	.05435	.34285
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.01800**	**.03876**	F **49.426**
Stddev	.00023	.00022	.593
%RSD	1.2842	.57319	1.2005
#1	.01823	.03881	50.072
#2	.01777	.03895	48.906
#3	.01800	.03851	49.300
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**7011.3**	**43477.**	**2210.1**
Stddev	20.4	40.	14.4
%RSD	.29146	.09218	.65240
#1	7004.2	43507.	2200.5
#2	6995.4	43491.	2226.7
#3	7034.3	43431.	2203.1

Sample Name: PBW 11 Acquired: 6/23/2016 22:35:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00032**	**.00826**	**.00073**	**.00097**	**-.00161**	**.00009**	**.02504**
Stddev	.00435	.00626	.00116	.00380	.00132	.00015	.01704
%RSD	1361.2	75.725	158.33	393.95	82.078	159.17	68.056
#1	-.00441	.01091	-.00035	.00280	-.00305	.00000	.03508
#2	.00122	.01276	.00059	.00350	-.00132	.00026	.03466
#3	.00414	.00112	.00196	-.00341	-.00045	.00001	.00536
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00010**	**-.00009**	**.00011**	**.00179**	**-.00443**	**.01747**	**-.00078**
Stddev	.00037	.00030	.00107	.00110	.02744	.11206	.00643
%RSD	372.56	345.12	981.79	61.211	619.83	641.53	826.74
#1	.00053	-.00038	.00079	.00288	.01410	-.01578	-.00199
#2	-.00009	.00021	-.00112	.00069	-.03595	.14238	.00617
#3	-.00014	-.00009	.00066	.00181	.00857	-.07420	-.00651
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.12116**	**.00145**	**-.00010**	**-.00638**	**-.00006**	**-.00844**	**-.00156**
Stddev	.07020	.00190	.00047	.05241	.00004	.00883	.00308
%RSD	57.940	131.08	475.47	821.85	65.273	104.67	197.22
#1	.10625	-.00074	-.00051	-.02102	-.00009	.00126	.00107
#2	.19761	.00240	-.00019	.05180	-.00002	-.01056	-.00495
#3	.05961	.00268	.00041	-.04991	-.00007	-.01601	-.00081
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: PBW 11 Acquired: 6/23/2016 22:35:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-02

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00235**	F **-.01321**	**-.01211**	**.00005**	**.00012**	**-.01090**	**-.00481**
Stddev	.00434	.00615	.00423	.00130	.00022	.00717	.00224
%RSD	184.96	46.568	34.938	2403.9	182.36	65.767	46.622
#1	-.00650	-.00846	-.00872	-.00069	-.00007	-.00262	-.00253
#2	.00215	-.02016	-.01076	-.00070	.00007	-.01519	-.00488
#3	-.00268	-.01101	-.01685	.00155	.00036	-.01489	-.00701
Check ?	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit		90.000					
Low Limit		-.01000					

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00056**	**.00115**	F **-.06214**
Stddev	.00033	.00011	.40590
%RSD	58.758	9.4676	653.23
#1	-.00093	.00115	.04499
#2	-.00040	.00127	.27945
#3	-.00033	.00105	-.51085
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6526.0**	**40939.**	**2061.7**
Stddev	35.4	294.	34.1
%RSD	.54178	.71898	1.6547
#1	6541.9	40985.	2026.5
#2	6550.5	40624.	2063.9
#3	6485.4	41207.	2094.6

Sample Name: LCSW 11 Acquired: 6/23/2016 22:38:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20572	5.1016	.20434	.99099	.52273	.02551	5.4134	.02557
Stddev	.00209	.0119	.00205	.00539	.00783	.00018	.0860	.00030
%RSD	1.0158	.23413	1.0041	.54353	1.4983	.70326	1.5893	1.1859
#1	.20812	5.1093	.20580	.99664	.52935	.02536	5.5127	.02591
#2	.20432	5.1075	.20521	.99043	.52477	.02571	5.3662	.02536
#3	.20472	5.0878	.20199	.98591	.51409	.02545	5.3613	.02542
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10385	.25890	.26189	2.1195	25.823	.53451	5.3645	.26403
Stddev	.00010	.00080	.00098	.0266	.322	.00518	.0615	.00364
%RSD	.09428	.30942	.37489	1.2540	1.2452	.96918	1.1457	1.3773
#1	.10374	.25918	.26254	2.1199	26.125	.53986	5.3994	.26710
#2	.10385	.25953	.26237	2.1460	25.858	.53416	5.2935	.26002
#3	.10394	.25800	.26076	2.0928	25.485	.52952	5.4005	.26498
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.53004	25.957	.26029	4.9232	.25825	.61920	.19466	2.6294
Stddev	.00026	.317	.00111	.0037	.00566	.00409	.00541	.0065
%RSD	.04851	1.2227	.42516	.07504	2.1932	.66079	2.7789	.24514
#1	.53022	26.268	.26156	4.9218	.25403	.62385	.18855	2.6314
#2	.53016	25.970	.25951	4.9204	.25605	.61617	.19656	2.6346
#3	.52975	25.634	.25981	4.9274	.26469	.61757	.19886	2.6222
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: LCSW 11 Acquired: 6/23/2016 22:38:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51706**	**.53317**	**.52168**	**.25600**	**.51571**	**.51503**	**1.3110**
Stddev	.00229	.00553	.01335	.00456	.00094	.00046	.2154
%RSD	.44200	1.0369	2.5594	1.7813	.18322	.08969	16.426
#1	.51564	.53801	.52874	.25386	.51673	.51503	1.4161
#2	.51585	.53436	.53002	.26124	.51555	.51550	1.4536
#3	.51970	.52714	.50628	.25291	.51486	.51457	1.0633
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6527.2**	**40455.**	**2087.1**
Stddev	31.3	163.	12.5
%RSD	.47880	.40351	.59802
#1	6563.2	40283.	2074.8
#2	6508.4	40474.	2086.8
#3	6509.8	40608.	2099.8

Sample Name: F BLANK Acquired: 6/23/2016 22:42:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573353-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00101	.01349	.00278	.00644	-.00048	.00013	.02829	-.00023
Stddev	.00147	.00646	.00238	.00275	.00027	.00015	.01012	.00024
%RSD	145.82	47.931	85.631	42.748	55.833	119.23	35.758	102.61
#1	-.00010	.00837	.00376	.00945	-.00079	.00030	.03838	-.00026
#2	.00267	.02075	.00451	.00583	-.00040	.00003	.02835	-.00045
#3	.00045	.01134	.00007	.00405	-.00027	.00005	.01815	.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00017	-.00021	.00115	.01618	-.02636	.00022	-.03112	.00301
Stddev	.00025	.00041	.00050	.01401	.02024	.00227	.10093	.00056
%RSD	146.69	193.40	43.311	86.581	76.800	1041.8	324.28	18.733
#1	.00022	.00012	.00058	.01388	-.01342	.00194	.07257	.00362
#2	.00039	-.00067	.00140	.00346	-.01597	.00107	-.03691	.00250
#3	-.00010	-.00008	.00148	.03119	-.04969	-.00236	-.12903	.00293
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00043	131.96	-.00024	-.00817	-.00215	.00148	-.00969	-.00282
Stddev	.00010	.85	.00144	.01034	.00162	.00392	.00917	.00327
%RSD	22.241	.64765	607.00	126.51	75.376	266.02	94.572	115.83
#1	.00044	132.92	.00143	.00262	-.00398	.00318	-.01589	.00093
#2	.00053	131.28	-.00110	-.01799	-.00092	.00426	.00084	-.00431
#3	.00034	131.69	-.00104	-.00915	-.00154	-.00301	-.01402	-.00509
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: F BLANK Acquired: 6/23/2016 22:42:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573353-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00016**	**.00010**	**-.00200**	**-.00314**	**.00056**	**.00410**	**.11736**
Stddev	.00046	.00052	.00261	.00021	.00036	.00009	.48090
%RSD	288.96	535.97	130.24	6.7340	64.659	2.2677	409.76
#1	-.00023	.00063	-.00290	-.00292	.00037	.00404	.00662
#2	.00004	.00007	-.00404	-.00316	.00033	.00421	-.29851
#3	.00067	-.00041	.00094	-.00334	.00097	.00406	.64398
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6547.2**	**40321.**	**2096.8**
Stddev	27.3	101.	33.4
%RSD	.41693	.25042	1.5916
#1	6550.7	40326.	2062.9
#2	6572.6	40217.	2129.6
#3	6518.3	40419.	2098.0

Sample Name: F BLANK Acquired: 6/23/2016 22:46:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573353-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00205**	**.01458**	**.00104**	**.00093**	**-.00012**	**-.00011**	**.03413**	**.00007**
Stddev	.00257	.00695	.00149	.00158	.00097	.00015	.06625	.00024
%RSD	125.57	47.660	142.55	168.73	807.82	139.24	194.10	358.50
#1	.00306	.01748	.00276	.00220	-.00059	.00003	.00736	.00006
#2	.00395	.00665	.00032	-.00083	-.00077	-.00008	-.01455	-.00017
#3	-.00087	.01960	.00006	.00144	.00099	-.00027	.10958	.00031
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**-.00025**	**.00229**	**.01454**	**.10361**	**-.00161**	**-.00074**	**.00159**
Stddev	.00020	.00099	.00086	.02039	.12854	.00327	.04972	.00150
%RSD	270.27	392.29	37.466	140.25	124.05	203.19	6704.2	94.513
#1	-.00016	-.00139	.00199	-.00883	.15515	.00182	-.05789	.00331
#2	.00020	.00021	.00326	.02371	.19839	-.00196	.02308	.00082
#3	.00019	.00042	.00163	.02874	-.04269	-.00469	.03259	.00062
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00013**	**.08993**	**.00021**	**-.00044**	**.00171**	**.00115**	**-.00478**	**-.01618**
Stddev	.00021	.02749	.00122	.00310	.00251	.00690	.00614	.00197
%RSD	158.62	30.569	571.20	708.88	146.77	601.02	128.46	12.198
#1	.00007	.10546	-.00067	.00234	-.00022	.00773	-.00669	-.01496
#2	.00036	.10614	-.00029	.00012	.00081	.00175	.00209	-.01845
#3	-.00004	.05819	.00160	-.00377	.00454	-.00604	-.00974	-.01512
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: F BLANK Acquired: 6/23/2016 22:46:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573353-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	.00000	-.00849	-.00355	.00002	.00375	.18786
Stddev	.00024	.00006	.00318	.00112	.00121	.00009	.14076
%RSD	171.57	4326.9	37.397	31.641	7603.4	2.4757	74.927
#1	.00012	.00006	-.01213	-.00271	-.00043	.00385	.34927
#2	.00039	-.00000	-.00626	-.00312	-.00090	.00367	.12376
#3	-.00009	-.00006	-.00708	-.00483	.00138	.00373	.09057
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6645.1	41415.	2118.1
Stddev	51.0	148.	9.8
%RSD	.76790	.35755	.46386
#1	6671.8	41525.	2108.0
#2	6677.3	41246.	2127.6
#3	6586.3	41473.	2118.7

Sample Name: L1606097401 Acquired: 6/23/2016 22:50:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00187**	**.11141**	**.00185**	**.13976**	**.13124**	**.00006**	**11.923**	**.00025**
Stddev	.00099	.00961	.00433	.00304	.00120	.00013	.085	.00026
%RSD	52.908	8.6272	234.23	2.1718	.91402	206.81	.71186	103.09
#1	-.00300	.11755	.00322	.14250	.13178	.00021	12.021	.00001
#2	-.00137	.11635	-.00300	.14028	.13207	-.00003	11.884	.00021
#3	-.00123	.10033	.00533	.13649	.12986	.00001	11.865	.00052
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00325**	**.00082**	**.00243**	**1.0194**	**3.4016**	**.00412**	**14.829**	**.47953**
Stddev	.00065	.00120	.00106	.0300	.0185	.00138	.053	.00284
%RSD	19.954	146.37	43.821	2.9431	.54459	33.491	.35710	.59200
#1	.00322	-.00007	.00334	.98836	3.3804	.00531	14.788	.47741
#2	.00262	.00219	.00268	1.0217	3.4096	.00444	14.889	.48276
#3	.00391	.00035	.00126	1.0482	3.4148	.00261	14.810	.47842
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00016**	**14.029**	**.00345**	**-.00115**	**.00096**	**.00344**	**.00640**	**6.1973**
Stddev	.00012	.046	.00111	.00836	.00159	.00074	.00619	.0079
%RSD	73.635	.32793	32.058	726.14	165.09	21.404	96.675	.12766
#1	-.00003	14.048	.00367	-.00743	.00071	.00392	.01206	6.2055
#2	-.00020	14.062	.00443	-.00436	-.00049	.00379	-.00020	6.1968
#3	-.00025	13.976	.00225	.00834	.00266	.00259	.00734	6.1897
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606097401 Acquired: 6/23/2016 22:50:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00010	.15663	-.00490	-.00350	-.00004	.00381	.27764
Stddev	.00040	.00118	.00403	.00013	.00036	.00005	.25173
%RSD	389.62	.75413	82.244	3.6390	976.32	1.2951	90.667
#1	-.00036	.15760	-.00950	-.00338	-.00044	.00385	.40246
#2	.00031	.15698	-.00318	-.00349	.00013	.00382	.44257
#3	.00036	.15532	-.00201	-.00363	.00021	.00375	-.01211
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6619.1	40945.	2109.7
Stddev	10.2	451.	19.0
%RSD	.15337	1.1019	.89902
#1	6615.2	40453.	2097.8
#2	6611.5	41044.	2131.5
#3	6630.6	41339.	2099.6

Sample Name: L1606097409S Acquired: 6/23/2016 22:54:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20627	5.0015	.19707	1.0840	.63830	.02511	15.216	.02489
Stddev	.00439	.0470	.00646	.0138	.00496	.00023	.159	.00037
%RSD	2.1297	.93911	3.2784	1.2764	.77719	.91601	1.0457	1.4914
#1	.20993	5.0515	.18974	1.0963	.64340	.02535	15.397	.02529
#2	.20140	4.9583	.19954	1.0690	.63801	.02510	15.151	.02456
#3	.20748	4.9946	.20193	1.0866	.63349	.02489	15.099	.02483
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10388	.25321	.25314	2.5849	28.688	.52179	17.865	.68847
Stddev	.00035	.00232	.00112	.0115	.394	.00543	.201	.00731
%RSD	.33918	.91513	.44229	.44279	1.3724	1.0405	1.1266	1.0613
#1	.10359	.25588	.25333	2.5862	29.131	.52794	18.088	.69576
#2	.10427	.25175	.25416	2.5957	28.557	.51978	17.697	.68851
#3	.10377	.25199	.25194	2.5729	28.377	.51766	17.810	.68114
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.51876	38.405	.25390	4.8710	.24934	.59995	.18589	8.2835
Stddev	.00146	.259	.00011	.0108	.00345	.00191	.00207	.0089
%RSD	.28048	.67387	.04392	.22193	1.3854	.31868	1.1116	.10799
#1	.51718	38.681	.25401	4.8591	.24740	.60185	.18570	8.2732
#2	.51906	38.367	.25379	4.8738	.25333	.59803	.18393	8.2891
#3	.52005	38.168	.25392	4.8802	.24730	.59997	.18804	8.2882
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606097409S Acquired: 6/23/2016 22:54:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50349	.65950	.51762	.24758	.50743	.50246	.81618
Stddev	.00174	.00452	.01522	.00062	.00357	.00056	.23521
%RSD	.34477	.68510	2.9397	.25174	.70344	.11237	28.818
#1	.50344	.66471	.52198	.24830	.51109	.50181	.58167
#2	.50178	.65703	.53018	.24729	.50396	.50287	.81481
#3	.50525	.65675	.50070	.24716	.50723	.50268	1.0521
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6571.1	40649.	2099.3
Stddev	19.5	291.	11.1
%RSD	.29604	.71659	.53063
#1	6548.6	40317.	2086.9
#2	6582.6	40767.	2102.9
#3	6582.0	40862.	2108.3

Sample Name: L1606097411SD Acquired: 6/23/2016 22:58:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20301	5.0935	.20254	1.1135	.64720	.02513	17.005	.02529
Stddev	.00131	.0093	.00255	.0073	.00193	.00014	.125	.00014
%RSD	.64434	.18208	1.2591	.65223	.29801	.54856	.73583	.55871
#1	.20180	5.0830	.20062	1.1056	.64841	.02524	16.944	.02527
#2	.20284	5.0968	.20544	1.1153	.64821	.02498	17.149	.02543
#3	.20440	5.1006	.20156	1.1198	.64498	.02518	16.923	.02515
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10473	.25680	.25656	3.0216	28.787	.52889	19.740	.73225
Stddev	.00029	.00169	.00156	.0240	.291	.00510	.166	.00214
%RSD	.27243	.65998	.60790	.79299	1.0111	.96496	.83971	.29254
#1	.10453	.25746	.25836	3.0492	28.796	.52755	19.834	.73428
#2	.10462	.25806	.25556	3.0077	29.074	.53453	19.837	.73246
#3	.10506	.25487	.25576	3.0078	28.492	.52458	19.548	.73001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.52618	39.581	.25576	4.8994	.25448	.60988	.19462	8.7250
Stddev	.00060	.184	.00118	.0146	.00431	.00152	.00488	.0127
%RSD	.11470	.46575	.46133	.29829	1.6939	.24983	2.5089	.14588
#1	.52623	39.585	.25707	4.8974	.25909	.60949	.19392	8.7122
#2	.52676	39.763	.25543	4.9148	.25381	.60859	.19982	8.7250
#3	.52555	39.395	.25478	4.8858	.25055	.61156	.19013	8.7376
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606097411SD Acquired: 6/23/2016 22:58:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573562-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50655	.68066	.51426	.25122	.50847	.50678	.50917
Stddev	.00036	.00376	.00731	.00121	.00054	.00037	.18705
%RSD	.07012	.55173	1.4206	.48320	.10539	.07293	36.736
#1	.50668	.67977	.51313	.25144	.50835	.50698	.47043
#2	.50615	.68478	.52206	.25231	.50800	.50700	.34452
#3	.50682	.67742	.50758	.24991	.50905	.50635	.71255
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6599.6	40879.	2125.2
Stddev	36.5	163.	36.7
%RSD	.55319	.39917	1.7271
#1	6558.7	40764.	2124.1
#2	6628.9	40808.	2089.1
#3	6611.1	41066.	2162.5

Sample Name: L1606090301 Acquired: 6/23/2016 23:01:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00171	.01564	.00294	.01850	.00898	-.00004	46.441	.00022
Stddev	.00082	.01258	.00166	.00390	.00087	.00003	.145	.00037
%RSD	47.964	80.448	56.393	21.081	9.6500	63.411	.31216	167.83
#1	.00264	.02660	.00383	.02298	.00987	-.00008	46.595	-.00020
#2	.00144	.01843	.00103	.01664	.00813	-.00002	46.423	.00039
#3	.00106	.00190	.00395	.01587	.00895	-.00004	46.307	.00049
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00038	.00043	.00251	.04392	.36940	-.00187	6.4848	.00401
Stddev	.00055	.00082	.00124	.02745	.02893	.00702	.0718	.00293
%RSD	143.61	190.56	49.243	62.491	7.8320	374.64	1.1071	73.244
#1	.00082	-.00015	.00183	.02878	.35514	-.00147	6.5673	.00719
#2	.00055	.00137	.00394	.07561	.35036	.00493	6.4364	.00342
#3	-.00023	.00007	.00177	.02738	.40269	-.00908	6.4507	.00141
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00124	2.3912	.00015	-.00111	.00175	-.00069	-.00611	2.4464
Stddev	.00019	.0146	.00065	.00585	.00193	.00258	.00753	.0043
%RSD	15.055	.61183	443.31	524.60	110.31	375.66	123.28	.17481
#1	.00123	2.3744	.00089	-.00577	.00397	.00193	-.00911	2.4476
#2	.00106	2.4012	-.00016	.00545	.00055	-.00077	.00246	2.4417
#3	.00143	2.3980	-.00029	-.00303	.00072	-.00322	-.01167	2.4500
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090301 Acquired: 6/23/2016 23:01:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00033	.05094	-.00942	-.00137	.00033	.00855	.12749
Stddev	.00120	.00036	.00450	.00125	.00064	.00008	.22777
%RSD	359.04	.70893	47.793	91.344	191.76	.98957	178.66
#1	-.00095	.05113	-.01374	-.00053	.00015	.00845	.36891
#2	.00142	.05116	-.00475	-.00281	-.00019	.00858	.09713
#3	.00053	.05052	-.00978	-.00077	.00105	.00861	-.08358
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6624.9	41305.	2113.7
Stddev	28.0	611.	12.8
%RSD	.42196	1.4799	.60649
#1	6599.0	41779.	2107.8
#2	6654.5	40615.	2128.4
#3	6621.1	41521.	2104.9

Sample Name: L1606090301PS Acquired: 6/23/2016 23:05:49 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573749-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20315	5.0351	.19933	.98509	.51723	.02533	46.108	.02517
Stddev	.00237	.0141	.00064	.00471	.00225	.00014	.292	.00010
%RSD	1.1667	.28092	.32345	.47796	.43513	.56006	.63289	.38937
#1	.20113	5.0207	.19918	.98035	.51851	.02535	46.344	.02507
#2	.20576	5.0490	.19878	.98516	.51463	.02547	45.781	.02527
#3	.20255	5.0357	.20004	.98977	.51856	.02518	46.198	.02516
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.10028	.25408	.25247	2.0837	25.422	.52185	10.797	.26020
Stddev	.00048	.00013	.00281	.0115	.145	.00302	.049	.00165
%RSD	.47862	.05151	1.1142	.55075	.56995	.57813	.44947	.63512
#1	.10024	.25422	.24943	2.0968	25.589	.52416	10.827	.25832
#2	.10077	.25406	.25499	2.0788	25.328	.51844	10.822	.26141
#3	.09982	.25396	.25300	2.0754	25.349	.52296	10.741	.26088
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.52036	27.383	.25017	4.9546	.25694	.60707	.19089	4.8076
Stddev	.00073	.158	.00061	.0081	.00347	.00719	.00485	.0007
%RSD	.13971	.57670	.24497	.16395	1.3493	1.1850	2.5386	.01435
#1	.52119	27.447	.24958	4.9632	.25410	.61287	.18545	4.8079
#2	.51984	27.204	.25012	4.9471	.25592	.60931	.19246	4.8068
#3	.52004	27.500	.25080	4.9535	.26081	.59902	.19476	4.8080
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090301PS Acquired: 6/23/2016 23:05:49 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573749-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50644	.56158	.50038	.25206	.50977	.50584	.50366
Stddev	.00169	.00120	.00839	.00343	.00237	.00060	.30226
%RSD	.33432	.21309	1.6764	1.3594	.46441	.11817	60.013
#1	.50760	.56240	.50354	.25552	.50926	.50626	.73300
#2	.50722	.56020	.49087	.24866	.51235	.50516	.61683
#3	.50450	.56212	.50672	.25200	.50770	.50611	.16114
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6634.2	40772.	2114.5
Stddev	22.4	92.	10.2
%RSD	.33771	.22603	.48307
#1	6611.3	40856.	2104.9
#2	6656.0	40786.	2113.4
#3	6635.2	40673.	2125.2

Sample Name: L1606090301SDL Acquired: 6/23/2016 23:09:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573749-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00125**	**-.00211**	**.00104**	**.00614**	**.00089**	**-.00005**	**8.6724**	**-.00005**
Stddev	.00276	.00609	.00315	.00512	.00101	.00009	.0375	.00023
%RSD	221.14	289.17	302.21	83.373	113.85	164.72	.43247	430.30
#1	.00165	-.00654	-.00255	.00543	.00189	-.00005	8.6376	.00022
#2	.00379	.00484	.00330	.01158	.00093	-.00014	8.6675	-.00019
#3	-.00169	-.00462	.00238	.00141	-.00014	.00003	8.7121	-.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00007**	**-.00015**	**.00151**	**.01774**	**.05719**	**-.00032**	**1.2928**	**.00199**
Stddev	.00032	.00036	.00174	.01658	.07315	.00592	.1368	.00072
%RSD	433.89	237.55	115.25	93.432	127.89	1849.8	10.583	36.311
#1	-.00025	-.00023	.00291	.03261	.10693	.00122	1.1631	.00203
#2	.00038	-.00046	-.00044	-.00013	-.02679	.00468	1.4358	.00125
#3	.00008	.00024	.00207	.02075	.09145	-.00685	1.2797	.00269
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00062**	**.50288**	**.00020**	**.00422**	**-.00347**	**-.00009**	**-.00488**	**.45851**
Stddev	.00015	.01796	.00107	.00310	.00056	.00216	.00737	.00307
%RSD	24.656	3.5717	530.84	73.548	16.155	2299.3	151.12	.67016
#1	.00074	.48937	-.00102	.00065	-.00411	-.00250	-.01008	.45511
#2	.00045	.49600	.00097	.00624	-.00312	.00057	.00356	.46107
#3	.00067	.52326	.00066	.00577	-.00317	.00165	-.00811	.45936
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090301SDL Acquired: 6/23/2016 23:09:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573749-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00043**	**.00975**	**-.00339**	**.00072**	**-.00071**	**.00182**	**.28982**
Stddev	.00076	.00047	.00412	.00137	.00059	.00016	.40983
%RSD	176.59	4.8404	121.60	190.63	83.639	8.7152	141.41
#1	.00038	.00960	.00073	.00044	-.00028	.00190	.42557
#2	-.00111	.01028	-.00752	.00221	-.00047	.00164	.61455
#3	-.00055	.00937	-.00339	-.00049	-.00139	.00193	-.17066
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6811.4**	**42010.**	**2145.9**
Stddev	35.6	211.	19.5
%RSD	.52290	.50301	.90902
#1	6804.9	42193.	2150.1
#2	6779.4	42057.	2163.0
#3	6849.8	41779.	2124.7

Sample Name: CCV Acquired: 6/23/2016 23:13:27 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41326**	**10.290**	**.40933**	**.51318**	**1.0406**	**.05196**	**10.449**
Stddev	.00367	.031	.00303	.00454	.0017	.00056	.073
%RSD	.88729	.29649	.74071	.88463	.16560	1.0752	.70056
#1	.41222	10.266	.40583	.51598	1.0426	.05202	10.433
#2	.41023	10.281	.41097	.50794	1.0395	.05138	10.384
#3	.41734	10.325	.41118	.51561	1.0396	.05249	10.528
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05188**	**.20751**	**.51860**	**.52298**	**4.1404**	**51.810**	**1.0557**
Stddev	.00032	.00026	.00373	.00169	.0293	.081	.0044
%RSD	.62111	.12412	.72010	.32358	.70759	.15585	.41608
#1	.05162	.20739	.51967	.52267	4.1341	51.858	1.0535
#2	.05178	.20781	.51444	.52146	4.1148	51.717	1.0529
#3	.05224	.20734	.52168	.52480	4.1724	51.855	1.0608
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**10.434**	**.52450**	**1.0396**	**51.672**	**.51702**	**10.257**	**.51943**
Stddev	.147	.00403	.0017	.210	.00153	.061	.00236
%RSD	1.4125	.76876	.16574	.40642	.29546	.59705	.45451
#1	10.599	.52887	1.0415	51.851	.51530	10.207	.52009
#2	10.315	.52369	1.0381	51.441	.51819	10.240	.51680
#3	10.387	.52093	1.0392	51.724	.51759	10.325	.52138
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 23:13:27 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2432**	**.40702**	**5.1978**	**1.0337**	**1.0394**	**1.0354**	**.52041**
Stddev	.0039	.00046	.0143	.0031	.0042	.0035	.00214
%RSD	.31258	.11246	.27417	.30049	.40463	.33586	.41209
#1	1.2391	.40753	5.1883	1.0302	1.0436	1.0350	.52283
#2	1.2437	.40689	5.1909	1.0348	1.0352	1.0322	.51875
#3	1.2468	.40665	5.2142	1.0361	1.0396	1.0391	.51965
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0351**	**1.0403**	F **.25266**
Stddev	.0070	.0016	.10822
%RSD	.68141	.15403	42.834
#1	1.0383	1.0386	.34911
#2	1.0270	1.0404	.13562
#3	1.0400	1.0418	.27323
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6474.0**	**39804.**	**2056.5**
Stddev	41.9	151.	9.5
%RSD	.64647	.37833	.46144
#1	6447.9	39827.	2053.8
#2	6522.3	39942.	2048.6
#3	6451.9	39644.	2067.0

Sample Name: CCB Acquired: 6/23/2016 23:17:05 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00053**	**.01688**	**.00315**	**.00457**	**-.00011**	**.00013**	**.01891**	**.00020**
Stddev	.00350	.01032	.00748	.00212	.00110	.00006	.06314	.00005
%RSD	664.12	61.108	237.21	46.535	973.24	44.359	333.92	24.066
#1	-.00268	.02858	-.00372	.00470	.00043	.00018	.03807	.00024
#2	-.00001	.00908	.00206	.00662	.00061	.00007	-.05159	.00021
#3	.00427	.01299	.01112	.00238	-.00138	.00016	.07025	.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00005**	**.00073**	**.00146**	**.01871**	**.09919**	**.00074**	**.03193**	**.00172**
Stddev	.00032	.00177	.00126	.01232	.12450	.00566	.07256	.00242
%RSD	651.38	241.69	86.427	65.838	125.51	769.76	227.22	140.80
#1	.00042	-.00122	.00034	.00498	-.03300	-.00467	.10382	.00139
#2	-.00019	.00221	.00282	.02879	.11637	.00026	-.04127	-.00052
#3	-.00008	.00121	.00121	.02235	.21421	.00662	.03325	.00429
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00396**	**.05479**	**-.00105**	**.00648**	**-.00101**	**.00564**	**-.00498**	**.00630**
Stddev	.00092	.03837	.00180	.00453	.00307	.00402	.00631	.00043
%RSD	23.350	70.024	171.44	69.981	304.27	71.194	126.78	6.8811
#1	.00290	.07487	.00083	.01030	.00189	.00526	-.01221	.00676
#2	.00456	.01055	-.00276	.00766	-.00422	.00984	-.00057	.00622
#3	.00443	.07894	-.00122	.00147	-.00069	.00183	-.00216	.00591
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: CCB Acquired: 6/23/2016 23:17:05 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00055**	**-.00002**	**.00408**	**-.00193**	**.00037**	**.00005**	**.02228**
Stddev	.00023	.00019	.00879	.00186	.00041	.00007	.40623
%RSD	41.382	1063.2	215.26	96.532	110.70	135.03	1823.1
#1	.00069	-.00023	-.00341	-.00071	-.00006	-.00002	-.40274
#2	.00029	.00004	.01375	-.00101	.00043	.00013	.06292
#3	.00066	.00013	.00190	-.00407	.00076	.00005	.40667
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6464.5**	**39816.**	**2046.0**
Stddev	26.9	481.	41.8
%RSD	.41677	1.2077	2.0425
#1	6458.5	39261.	2010.7
#2	6441.1	40106.	2035.2
#3	6493.9	40080.	2092.2

Sample Name: L1606090302 Acquired: 6/23/2016 23:21:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00031	.00800	.00182	.01230	.01036	-.00005	47.673	.00026
Stddev	.00085	.00343	.00363	.00358	.00124	.00007	.280	.00015
%RSD	272.61	42.881	200.18	29.113	11.960	128.20	.58740	57.099
#1	.00031	.00879	-.00237	.00886	.01157	-.00001	47.890	.00037
#2	-.00053	.01096	.00415	.01601	.00909	-.00001	47.772	.00032
#3	.00116	.00424	.00367	.01205	.01042	-.00013	47.357	.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00005	.00015	.00261	.09179	.46952	-.00272	6.7874	.01530
Stddev	.00029	.00088	.00094	.00788	.11796	.00383	.0475	.00163
%RSD	527.39	592.21	36.175	8.5807	25.123	140.65	.69987	10.685
#1	-.00038	.00116	.00159	.08468	.47413	-.00254	6.7436	.01564
#2	.00012	-.00040	.00278	.09044	.34932	.00101	6.7806	.01352
#3	.00010	-.00032	.00345	.10026	.58509	-.00664	6.8379	.01673
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00092	2.5218	-.00025	.00306	.00002	-.00071	-.00373	2.5222
Stddev	.00063	.0302	.00136	.00470	.00238	.00277	.00475	.0109
%RSD	68.779	1.1957	538.26	153.23	11843.	388.81	127.31	.43311
#1	.00032	2.5566	.00100	-.00086	.00039	.00133	.00175	2.5345
#2	.00086	2.5031	-.00169	.00178	.00219	.00040	-.00620	2.5136
#3	.00157	2.5056	-.00007	.00827	-.00252	-.00386	-.00675	2.5186
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090302 Acquired: 6/23/2016 23:21:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00028**	**.05280**	**-.00935**	**-.00377**	**-.00005**	**.00235**	**.11164**
Stddev	.00108	.00027	.00453	.00048	.00146	.00015	.26019
%RSD	392.25	.51917	48.460	12.830	2892.9	6.5370	233.05
#1	-.00049	.05310	-.00574	-.00424	.00086	.00248	.33955
#2	.00090	.05256	-.01444	-.00328	.00072	.00218	-.17184
#3	-.00123	.05274	-.00787	-.00379	-.00173	.00238	.16722
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6545.6**	**41303.**	**2131.1**
Stddev	23.7	132.	13.6
%RSD	.36137	.32048	.63580
#1	6530.1	41250.	2117.0
#2	6572.8	41206.	2132.3
#3	6534.0	41454.	2144.0

Sample Name: L1606090303 Acquired: 6/23/2016 23:25:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00263	.00774	.00563	.00796	.00854	-.00002	45.954	.00013
Stddev	.00328	.00668	.00265	.00318	.00150	.00012	.228	.00034
%RSD	124.84	86.299	47.034	39.934	17.503	682.86	.49687	267.94
#1	.00506	.01059	.00367	.00860	.00804	-.00005	45.886	.00049
#2	-.00111	.01253	.00457	.01077	.01022	.00012	46.208	.00008
#3	.00394	.00011	.00864	.00451	.00737	-.00012	45.767	-.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00022	-.00004	.00264	.21996	.44485	.00448	6.7971	.02189
Stddev	.00017	.00063	.00127	.02132	.06435	.00051	.0747	.00331
%RSD	77.701	1729.3	48.231	9.6943	14.465	11.277	1.0984	15.131
#1	.00007	.00062	.00404	.23944	.50211	.00394	6.7547	.01807
#2	.00019	-.00009	.00157	.19718	.45723	.00494	6.8833	.02378
#3	.00041	-.00064	.00230	.22325	.37521	.00457	6.7534	.02384
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00090	2.4891	-.00031	.00226	-.00318	.00121	-.00101	2.4462
Stddev	.00023	.0385	.00053	.00868	.00252	.00105	.00896	.0110
%RSD	25.268	1.5452	169.73	384.84	79.079	87.283	885.01	.44916
#1	.00089	2.4503	-.00059	-.00015	-.00309	-.00001	-.00785	2.4500
#2	.00067	2.4900	.00030	.01188	-.00575	.00184	-.00433	2.4549
#3	.00113	2.5272	-.00064	-.00497	-.00072	.00180	.00914	2.4339
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090303 Acquired: 6/23/2016 23:25:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.05042	-.00867	-.00405	-.00058	.00293	.21469
Stddev	.00068	.00039	.00321	.00395	.00086	.00012	.24147
%RSD	116.53	.76868	37.044	97.556	147.98	4.1433	112.48
#1	.00096	.05045	-.00617	-.00495	-.00048	.00287	.22391
#2	.00100	.05080	-.01229	-.00746	-.00148	.00307	.45141
#3	-.00020	.05002	-.00756	.00027	.00022	.00286	-.03127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6557.1	40614.	2090.8
Stddev	3.1	286.	4.7
%RSD	.04661	.70434	.22607
#1	6559.9	40616.	2093.3
#2	6553.9	40327.	2085.3
#3	6557.7	40899.	2093.7

Sample Name: L1606090304 Acquired: 6/23/2016 23:28:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00249	.01012	.00464	.01066	.00806	-.00002	46.697	.00023
Stddev	.00091	.01191	.00138	.00192	.00037	.00002	.216	.00013
%RSD	36.530	117.71	29.729	17.988	4.5683	108.87	.46335	55.814
#1	.00283	-.00243	.00466	.01137	.00780	-.00002	46.940	.00033
#2	.00319	.01152	.00601	.00849	.00791	.00000	46.527	.00026
#3	.00146	.02127	.00325	.01212	.00849	-.00003	46.623	.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00033	.00062	.00100	.51001	.64738	.00217	6.6353	.05303
Stddev	.00037	.00064	.00122	.03061	.09484	.00222	.1713	.00028
%RSD	109.48	103.02	122.22	6.0022	14.650	102.21	2.5823	.52039
#1	-.00001	.00028	.00000	.50309	.73158	.00219	6.5308	.05285
#2	-.00026	.00023	.00063	.48345	.66592	.00439	6.8331	.05291
#3	-.00073	.00136	.00236	.54349	.54463	-.00006	6.5421	.05335
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00050	2.5222	.00010	.00174	.00050	.00128	-.00532	2.7225
Stddev	.00078	.0513	.00045	.01103	.00453	.00439	.00901	.0008
%RSD	156.95	2.0349	434.03	634.47	913.70	343.72	169.31	.03118
#1	-.00033	2.5754	.00043	-.00461	.00243	-.00315	-.01182	2.7217
#2	.00121	2.4730	-.00041	.01448	.00374	.00135	-.00910	2.7234
#3	.00061	2.5183	.00029	-.00465	-.00468	.00562	.00496	2.7225
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090304 Acquired: 6/23/2016 23:28:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00081**	**.04960**	**-.00329**	**-.00196**	**-.00058**	**.00257**	**.19345**
Stddev	.00125	.00047	.00925	.00100	.00022	.00015	.14173
%RSD	153.83	.94383	281.36	50.890	37.305	5.7376	73.263
#1	-.00194	.04918	-.00253	-.00307	-.00034	.00241	.33632
#2	.00053	.04951	.00556	-.00167	-.00074	.00270	.19114
#3	-.00102	.05011	-.01289	-.00114	-.00067	.00259	.05289
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6594.2**	**40470.**	**2074.5**
Stddev	30.7	142.	46.3
%RSD	.46503	.35187	2.2319
#1	6626.4	40634.	2027.0
#2	6565.3	40392.	2077.0
#3	6590.8	40384.	2119.5

Sample Name: L1606090305 Acquired: 6/23/2016 23:32:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00213	.01717	-.00084	.01089	.00981	.00002	48.923	.00033
Stddev	.00269	.00400	.00320	.00082	.00051	.00009	.164	.00010
%RSD	126.20	23.278	380.73	7.5221	5.1474	403.64	.33444	29.068
#1	-.00080	.01450	.00248	.01170	.01029	.00012	49.107	.00031
#2	.00271	.02176	-.00390	.01090	.00928	.00001	48.795	.00044
#3	.00448	.01524	-.00110	.01007	.00986	-.00006	48.866	.00025
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00020	-.00068	.00177	1.5788	.44150	.00018	6.7726	.08864
Stddev	.00038	.00111	.00108	.0172	.06024	.00100	.1054	.00177
%RSD	191.06	162.17	60.718	1.0921	13.644	545.86	1.5565	1.9982
#1	-.00014	-.00131	.00085	1.5835	.49954	.00095	6.8808	.08930
#2	.00013	.00060	.00296	1.5597	.37928	.00054	6.7668	.08663
#3	.00060	-.00134	.00152	1.5932	.44569	-.00095	6.6702	.08998
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00035	2.4907	.00039	.00113	.00257	.00234	-.00366	3.0530
Stddev	.00062	.0095	.00077	.00213	.00167	.00385	.01135	.0045
%RSD	175.61	.38135	200.05	188.45	64.988	164.32	310.04	.14641
#1	-.00035	2.4923	-.00007	.00130	.00338	.00574	-.00469	3.0528
#2	.00078	2.4993	.00128	-.00108	.00370	-.00184	-.01446	3.0486
#3	.00062	2.4805	-.00005	.00317	.00065	.00313	.00817	3.0576
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: L1606090305 Acquired: 6/23/2016 23:32:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00063**	**.05089**	**-.00933**	**-.00366**	**-.00080**	**.01192**	**.09534**
Stddev	.00059	.00042	.00867	.00169	.00053	.00018	.35498
%RSD	93.514	.82669	92.914	46.068	66.583	1.5108	372.34
#1	-.00036	.05120	-.01547	-.00172	-.00063	.01174	-.06523
#2	-.00022	.05105	.00059	-.00451	-.00139	.01210	.50224
#3	-.00130	.05041	-.01311	-.00475	-.00037	.01192	-.15099
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6602.9**	**40942.**	**2135.9**
Stddev	6.9	324.	25.7
%RSD	.10425	.79054	1.2024
#1	6597.6	40623.	2114.4
#2	6610.7	41270.	2164.3
#3	6600.5	40933.	2129.0

Sample Name: L1606092401 Acquired: 6/23/2016 23:36:50 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00265	.03744	.60307	2.0253	.01194	.00006	59.280
Stddev	.00277	.00788	.00343	.0045	.00047	.00007	.270
%RSD	104.56	21.060	.56883	.22109	3.9165	122.30	.45465
#1	.00477	.04143	.60414	2.0202	.01227	.00013	59.521
#2	-.00048	.02835	.59924	2.0278	.01216	-.00002	58.989
#3	.00365	.04252	.60585	2.0280	.01141	.00007	59.330
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00037	.00169	.04543	.00223	.02086	20.304	.12271
Stddev	.00012	.00069	.00181	.00172	.03477	.327	.00387
%RSD	33.446	40.936	3.9734	76.929	166.67	1.6108	3.1576
#1	.00045	.00112	.04573	.00336	.05783	20.550	.12442
#2	.00044	.00245	.04349	.00026	.01592	20.429	.12545
#3	.00023	.00148	.04706	.00308	-.01117	19.933	.11828
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.44128	.00138	.09389	F 925.96	-.00083	.00321	-.00198
Stddev	.02883	.00107	.00073	20.38	.00037	.00049	.00073
%RSD	6.5329	77.799	.77502	2.2006	45.088	15.416	37.027
#1	.47443	.00015	.09321	948.64	-.00126	.00268	-.00137
#2	.42732	.00186	.09382	909.19	-.00057	.00328	-.00178
#3	.42208	.00212	.09466	920.06	-.00066	.00367	-.00280
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Sample Name: L1606092401 Acquired: 6/23/2016 23:36:50 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02428**	**.03016**	**5.4351**	**-.00127**	**.13763**	**-.01760**	**-.00342**
Stddev	.00750	.01129	.0055	.00172	.00098	.00201	.00308
%RSD	30.891	37.425	.10042	135.11	.71182	11.421	90.086
#1	.02439	.02726	5.4392	.00004	.13847	-.01734	-.00405
#2	.03173	.04262	5.4371	-.00064	.13656	-.01973	-.00008
#3	.01673	.02061	5.4289	-.00322	.13787	-.01573	-.00615
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.00453**	**.02225**	F **-.12201**
Stddev	.00050	.00015	.18874
%RSD	11.054	.67831	154.69
#1	.00443	.02208	-.21896
#2	.00408	.02238	-.24258
#3	.00507	.02228	.09550
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6296.4**	**38355.**	**2086.6**
Stddev	9.5	124.	22.4
%RSD	.15057	.32201	1.0746
#1	6287.1	38290.	2062.1
#2	6296.1	38498.	2106.2
#3	6306.0	38278.	2091.5

Sample Name: DI7 Acquired: 6/23/2016 23:40:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00122	-.00305	.00325	.01342	.00511	-.00003	.63457	.00006
Stddev	.00070	.00725	.00183	.00147	.00092	.00006	.03174	.00026
%RSD	57.204	237.69	56.132	10.964	18.001	169.74	5.0023	454.53
#1	.00129	-.01124	.00390	.01220	.00527	-.00006	.66854	.00019
#2	.00049	.00254	.00119	.01301	.00412	-.00007	.60566	-.00024
#3	.00189	-.00045	.00467	.01505	.00594	.00003	.62951	.00022
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00034	-.00105	.00137	.04887	.17552	-.00411	.67989	.02289
Stddev	.00029	.00066	.00135	.03758	.14542	.00295	.10949	.00017
%RSD	86.943	62.496	98.260	76.910	82.851	71.791	16.104	.73432
#1	.00007	-.00034	.00163	.08822	.25845	-.00273	.72653	.02297
#2	.00065	-.00118	.00257	.01334	.00761	-.00210	.75834	.02270
#3	.00029	-.00164	-.00009	.04504	.26051	-.00749	.55480	.02300
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00034	.88120	.00036	.00359	-.00382	-.00015	.00024	.27765
Stddev	.00027	.03865	.00120	.00516	.00377	.00137	.00372	.00507
%RSD	80.796	4.3866	328.28	143.80	98.655	910.66	1547.5	1.8265
#1	-.00036	.92576	.00056	.00038	-.00770	-.00061	-.00359	.27282
#2	-.00006	.85668	-.00092	.00954	-.00358	.00139	.00047	.27719
#3	-.00060	.86116	.00145	.00084	-.00017	-.00123	.00384	.28294
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: DI7 Acquired: 6/23/2016 23:40:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00014**	**.00711**	**-.00708**	**-.00329**	**-.00065**	**.00225**	**.10588**
Stddev	.00142	.00039	.01416	.00145	.00012	.00019	.39638
%RSD	991.52	5.5260	199.93	44.028	18.378	8.2791	374.37
#1	-.00086	.00703	-.01555	-.00371	-.00072	.00228	.22011
#2	.00150	.00677	.00926	-.00449	-.00071	.00242	-.33508
#3	-.00107	.00754	-.01496	-.00168	-.00051	.00205	.43261
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6937.5**	**42404.**	**2145.0**
Stddev	31.4	540.	24.8
%RSD	.45241	1.2743	1.1557
#1	6923.1	42362.	2162.5
#2	6915.9	42964.	2155.8
#3	6973.5	41886.	2116.6

Sample Name: DI8 Acquired: 6/23/2016 23:44:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01092	.24637	.01430	.05175	.02804	.00125	.70558	.00137
Stddev	.00255	.00988	.00240	.00085	.00181	.00008	.05906	.00031
%RSD	23.343	4.0097	16.798	1.6422	6.4611	6.1453	8.3704	22.665
#1	.01275	.23892	.01336	.05090	.02859	.00117	.73476	.00163
#2	.01199	.25758	.01250	.05173	.02602	.00133	.63761	.00146
#3	.00801	.24262	.01702	.05260	.02952	.00124	.74437	.00102
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00507	.01122	.01287	.11447	1.4023	.02306	.83494	.03245
Stddev	.00024	.00067	.00189	.02802	.1110	.00136	.03948	.00268
%RSD	4.7987	6.0062	14.707	24.480	7.9171	5.8811	4.7291	8.2701
#1	.00493	.01046	.01351	.14674	1.4178	.02155	.80587	.03520
#2	.00493	.01147	.01074	.10042	1.2843	.02345	.81905	.03233
#3	.00535	.01173	.01436	.09625	1.5047	.02418	.87989	.02983
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02127	1.8562	.01215	.23280	.01067	.03011	.00922	.37328
Stddev	.00053	.0342	.00120	.00909	.00269	.00363	.00276	.00458
%RSD	2.4934	1.8408	9.8651	3.9061	25.214	12.073	29.987	1.2271
#1	.02172	1.8278	.01327	.24205	.00898	.03426	.00781	.37038
#2	.02141	1.8468	.01231	.22387	.01377	.02752	.01240	.37091
#3	.02069	1.8941	.01088	.23248	.00925	.02853	.00745	.37856
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: DI8 Acquired: 6/23/2016 23:44:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02095**	**.02967**	**.01993**	**.00726**	**.02346**	**.02475**	**.30772**
Stddev	.00032	.00011	.00361	.00564	.00093	.00004	.15766
%RSD	1.5278	.37363	18.102	77.691	3.9489	.16512	51.234
#1	.02080	.02956	.01706	.01372	.02253	.02472	.37227
#2	.02072	.02978	.02398	.00329	.02439	.02479	.42286
#3	.02131	.02968	.01875	.00477	.02345	.02473	.12803
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6883.6**	**42542.**	**2132.1**
Stddev	35.9	81.	32.8
%RSD	.52157	.19075	1.5391
#1	6873.8	42453.	2096.2
#2	6923.4	42611.	2160.5
#3	6853.7	42564.	2139.6

Sample Name: DI9 Acquired: 6/23/2016 23:48:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.01172**	**.24362**	**.00993**	**.05401**	**.02815**	**.00127**	**.79730**
Stddev	.00152	.00569	.00166	.00158	.00083	.00011	.02246
%RSD	12.924	2.3335	16.753	2.9338	2.9467	8.5602	2.8175
#1	.01069	.24626	.01067	.05290	.02910	.00116	.79850
#2	.01102	.23710	.01110	.05583	.02755	.00137	.81914
#3	.01346	.24751	.00803	.05330	.02780	.00127	.77426
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00129**	**.00488**	**.01079**	**.01208**	**.12723**	**1.2786**	**.02511**
Stddev	.00009	.00006	.00031	.00168	.01475	.0943	.00185
%RSD	6.9948	1.2164	2.8343	13.930	11.594	7.3784	7.3599
#1	.00122	.00481	.01066	.01339	.14076	1.1744	.02529
#2	.00139	.00493	.01057	.01268	.11150	1.3582	.02686
#3	.00125	.00489	.01114	.01018	.12944	1.3031	.02318
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.89576**	**.03366**	**.02131**	**1.9045**	**.01104**	**.23277**	**.01075**
Stddev	.05721	.00137	.00010	.0400	.00057	.00507	.00231
%RSD	6.3863	4.0674	.47440	2.0992	5.1553	2.1800	21.454
#1	.95963	.03212	.02142	1.8638	.01067	.23321	.01329
#2	.84922	.03409	.02122	1.9059	.01170	.22750	.00878
#3	.87844	.03476	.02130	1.9437	.01076	.23762	.01018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: DI9 Acquired: 6/23/2016 23:48:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.03181**	**.00568**	**.38831**	**.02056**	**.03131**	**.02182**	**.01037**
Stddev	.00137	.01144	.00136	.00010	.00017	.00825	.00275
%RSD	4.3124	201.38	.34995	.48690	.53387	37.784	26.554
#1	.03338	.01685	.38970	.02064	.03120	.02474	.00719
#2	.03087	-.00602	.38824	.02045	.03123	.02822	.01188
#3	.03117	.00622	.38698	.02060	.03150	.01252	.01204
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.02300**	**.02452**	F **-.17091**
Stddev	.00069	.00018	.04679
%RSD	2.9886	.74679	27.377
#1	.02312	.02467	-.13285
#2	.02363	.02457	-.15673
#3	.02227	.02431	-.22315
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6863.0**	**42455.**	**2140.2**
Stddev	39.9	478.	31.6
%RSD	.58151	1.1260	1.4765
#1	6898.2	41962.	2103.8
#2	6819.6	42485.	2156.2
#3	6871.2	42917.	2160.6

Sample Name: CCV Acquired: 6/23/2016 23:52:52 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.41335	10.421	.41508	.51836	1.0496	.05256	10.492
Stddev	.00039	.045	.00377	.00553	.0024	.00034	.011
%RSD	.09500	.43207	.90936	1.0674	.22730	.64627	.10764
#1	.41368	10.369	.41539	.51316	1.0470	.05217	10.480
#2	.41292	10.443	.41116	.51776	1.0502	.05276	10.503
#3	.41347	10.451	.41869	.52417	1.0517	.05276	10.494
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05201	.20974	.52865	.52594	4.2034	52.520	1.0760
Stddev	.00016	.00127	.00164	.00103	.0566	.261	.0077
%RSD	.30575	.60470	.30961	.19638	1.3473	.49635	.71934
#1	.05219	.20874	.52729	.52653	4.2121	52.227	1.0702
#2	.05189	.20932	.52820	.52475	4.1429	52.726	1.0848
#3	.05194	.21117	.53047	.52654	4.2552	52.606	1.0729
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	10.516	.52754	1.0503	52.353	.52349	10.349	.52208
Stddev	.065	.00345	.0031	.035	.00255	.035	.00042
%RSD	.61612	.65474	.29626	.06645	.48722	.33829	.08130
#1	10.587	.52445	1.0535	52.313	.52122	10.327	.52177
#2	10.461	.53127	1.0472	52.379	.52299	10.332	.52191
#3	10.500	.52692	1.0503	52.366	.52625	10.390	.52256
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/23/2016 23:52:52 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2512**	**.41632**	**5.2448**	**1.0458**	**1.0501**	**1.0483**	**.52498**
Stddev	.0079	.00080	.0298	.0049	.0030	.0126	.00266
%RSD	.62850	.19294	.56895	.46593	.28810	1.2026	.50623
#1	1.2568	.41583	5.2243	1.0406	1.0479	1.0390	.52296
#2	1.2422	.41589	5.2311	1.0465	1.0488	1.0433	.52399
#3	1.2546	.41725	5.2791	1.0503	1.0535	1.0626	.52799
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0496**	**1.0495**	F **.09970**
Stddev	.0053	.0039	.10577
%RSD	.50966	.36953	106.08
#1	1.0434	1.0481	.19784
#2	1.0529	1.0466	.11360
#3	1.0524	1.0539	-.01233
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6372.6**	**39443.**	**2025.2**
Stddev	36.3	113.	32.7
%RSD	.56912	.28682	1.6135
#1	6379.4	39570.	1991.0
#2	6404.9	39404.	2028.4
#3	6333.4	39354.	2056.1

Sample Name: CCB Acquired: 6/23/2016 23:56:30 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00038**	**.01156**	**.00056**	**.00234**	**.00056**	**.00010**	**.03716**
Stddev	.00349	.01000	.00359	.00308	.00040	.00012	.04368
%RSD	913.43	86.498	638.46	132.02	72.010	128.40	117.56
#1	.00278	.00590	.00379	.00161	.00011	-.00003	.06293
#2	.00200	.02311	.00120	.00572	.00089	.00010	-.01328
#3	-.00362	.00567	-.00330	-.00032	.00068	.00022	.06182
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00024**	**.00016**	**.00067**	**.00127**	**.01578**	**.03402**	**-.00346**
Stddev	.00007	.00014	.00147	.00047	.02063	.10839	.00477
%RSD	29.479	83.436	221.45	36.697	130.75	318.63	137.81
#1	.00027	.00008	-.00033	.00181	.00506	.08712	-.00283
#2	.00029	.00009	-.00003	.00096	.00272	-.09068	-.00851
#3	.00016	.00032	.00236	.00106	.03957	.10561	.00096
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00827**	**.00236**	**.00382**	**.04976**	**.00076**	**.00558**	**-.00126**
Stddev	.03186	.00300	.00012	.02808	.00054	.00951	.00325
%RSD	385.40	127.33	3.0647	56.434	70.960	170.62	257.97
#1	-.01385	.00576	.00377	.02300	.00097	.01389	-.00238
#2	-.00614	.00008	.00395	.07899	.00015	-.00480	.00240
#3	.04479	.00123	.00373	.04728	.00116	.00763	-.00379
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: CCB Acquired: 6/23/2016 23:56:30 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00469**	**-.00110**	**.00540**	**-.00045**	**.00020**	**-.00579**	**-.00171**
Stddev	.00519	.00566	.00080	.00082	.00023	.00790	.00277
%RSD	110.60	514.68	14.743	182.74	110.97	136.58	162.08
#1	.00345	.00405	.00520	.00050	.00023	-.00735	-.00373
#2	.01039	-.00018	.00628	-.00094	.00042	-.01279	.00145
#3	.00024	-.00717	.00473	-.00090	-.00003	.00278	-.00284
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00035**	**.00012**	F **-.30075**
Stddev	.00143	.00011	.38164
%RSD	413.19	92.936	126.90
#1	.00023	.00022	-.11902
#2	.00071	-.00000	-.73929
#3	-.00198	.00015	-.04394
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			.04000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6443.5**	**39921.**	**2022.5**
Stddev	30.3	559.	26.8
%RSD	.46979	1.3996	1.3238
#1	6467.0	40295.	2021.5
#2	6409.3	40190.	1996.3
#3	6454.1	39279.	2049.8

Sample Name: ICSA Acquired: 6/24/2016 0:00:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	s -.00181	^F *****	k .00039	s .01077	F -16.124	s .00066
Stddev	.00405	-----	.00084	.01107	27.933	.00100
%RSD	224.07	-----	217.80	102.72	173.24	151.08
#1	s .00084	s .38723	k .00033	s .00228	-48.378	s -.00024
#2	s .00021	s 13.731	k .00125	s .00675	.0154	s .00048
#3	s -.00648	^ -----	k -.00043	s .02329	-.0093	s .00173
Check ?	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit		300.00			.00400	
Low Limit		200.00			-.00400	

Elem	Ca4226	Cd2288	Co2286	Cr2677	Cu2247	Fe2611
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F -17.795	kF -.00145	k .00006	s .00397	kF .00900	kF 13.080
Stddev	30.964	.00234	.00025	.00687	.00196	22.623
%RSD	174.00	161.85	431.98	173.11	21.777	172.95
#1	-53.549	k -.00415	k .00030	s -.00220	k .00711	k 39.203
#2	.1724	k .00000	k .00007	s .00273	k .00885	k .00481
#3	-.0087	k -.00020	k -.00020	s .01137	k .01102	k .03331
Check ?	Chk Fail	Chk Fail	Chk Pass	Chk Pass	Chk Fail	Chk Fail
High Limit	300.00	.00080			.00400	120.00
Low Limit	200.00	-.00080			-.00400	80.000

Elem	K_7664	Li6707	Mg2790	Mn2576	Mo2020	Na5895
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F 303.15	F 17.009	kF 144.99	kF -2.2313	k .00055	F 349.42
Stddev	524.72	29.469	251.19	3.8674	.00057	605.42
%RSD	173.09	173.26	173.25	173.32	104.51	173.26
#1	909.05	51.038	k 435.03	k -6.6970	k -.00009	1048.5
#2	.2954	-.01212	k -.0973	k .00125	k .00102	-.464
#3	.1091	.00231	k .0233	k .00181	k .00071	.231
Check ?	Chk Fail	Chk Fail	Chk Fail	Chk Fail	Chk Pass	Chk Fail
High Limit	.40000	.08000	300.00	.00400		.25000
Low Limit	-.40000	-.08000	200.00	-.00400		-.25000

Sample Name: ICSA Acquired: 6/24/2016 0:00:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	k **.00287**	k **.00325**	kF **.01305**	k **.00093**	k **-.00095**	k **-.02524**
Stddev	.00051	.00655	.00114	.00346	.00637	.00077
%RSD	17.929	201.56	8.7344	372.00	672.86	3.0546
#1	k .00291	k .00302	k .01321	k .00198	k -.00342	k -.02442
#2	k .00336	k .00992	k .01183	k .00374	k .00629	k -.02594
#3	k .00234	k -.00319	k .01409	k -.00293	k -.00571	k -.02536
Check ?	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit			.00400			
Low Limit			-.00400			

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	k **-.00038**	F **-.53580**	kF **27.418**	k **.00263**	s **-.00151**	k **.00080**
Stddev	.00064	.93031	47.540	.00108	.00198	.00005
%RSD	165.52	173.63	173.39	41.177	131.41	6.5158
#1	k -.00073	-1.6100	k 82.313	k .00161	s .00053	k .00074
#2	k -.00077	.00282	k -.07422	k .00377	s -.00163	k .00084
#3	k .00035	-.00020	k .01560	k .00250	s -.00343	k .00082
Check ?	Chk Pass	Chk Fail	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit		.00800	.02400			
Low Limit		-.00800	-.02400			

Elem	Zr3391
Units	ppm
Avg	kF **2038.2**
Stddev	3532.4
%RSD	173.31
#1	k 6117.0
#2	k -4.501
#3	k 2.059
Check ?	Chk Fail
High Limit	.02000
Low Limit	-.02000

Sample Name: ICSA Acquired: 6/24/2016 0:00:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9727.2**	^ *****	**2367.9**
Stddev	176.1	-----	3403.9
%RSD	1.8108	-----	143.75
#1	9699.2	^ -----	-.533
#2	9915.7	^ -----	835.60
#3	9566.8	^ -----	6268.6

Sample Name: ICSAB Acquired: 6/24/2016 0:04:29 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	sF .03510	^F *****	kF .02243	s -.00104	F -17.624	sF .01622
Stddev	.03890	-----	.00215	.00199	39.067	.01920
%RSD	110.83	-----	9.5662	190.62	221.67	118.34
#1	s .00126	s -.05314	k .01995	s -.00051	-62.399	s -.00008
#2	s .02644	^ -----	k .02376	s .00063	9.5264	s .01137
#3	s .07761	s 36.789	k .02358	s -.00324	-.0004	s .03738
Check ?	Chk Fail	Chk Fail	Chk Fail	Chk Pass	Chk Fail	Chk Fail
High Limit	.60000	300.00	.30000		.30000	.30000
Low Limit	.40000	200.00	.20000		.20000	.20000

Elem	Ca4226	Cd2288	Co2286	Cr2677	Cu2247	Fe2611
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F -41.394	kF .04368	kF .02300	sF .01591	kF .03181	kF -23.435
Stddev	80.483	.00655	.00163	.02148	.00204	29.861
%RSD	194.43	14.997	7.0968	135.04	6.3986	127.42
#1	-134.15	k .03613	k .02112	s -.00134	k .02989	k -57.047
#2	9.956	k .04777	k .02379	s .00909	k .03394	k -13.288
#3	.012	k .04715	k .02408	s .03996	k .03158	k .0307
Check ?	Chk Fail	Chk Fail	Chk Fail	Chk Fail	Chk Fail	Chk Fail
High Limit	300.00	.60000	.30000	.30000	.30000	120.00
Low Limit	200.00	.40000	.20000	.20000	.20000	80.000

Elem	K_7664	Li6707	Mg2790	Mn2576	Mo2020	Na5895
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F 64.851	F 12.009	kF -20.121	kF 2.2589	k .00017	F 379.32
Stddev	297.64	30.686	17.446	6.5094	.00008	850.62
%RSD	458.96	255.52	86.705	288.17	49.010	224.25
#1	389.61	46.885	k -30.037	k 9.5964	k .00008	1353.6
#2	-194.93	-10.851	k -30.348	k -2.8212	k .00024	-215.77
#3	-.126	-.0058	k .0229	k .00136	k .00019	.137
Check ?	Chk Fail	Chk Fail	Chk Fail	Chk Fail	Chk Pass	Chk Fail
High Limit	6.0000	.10000	300.00	.30000		6.0000
Low Limit	4.0000	-.10000	200.00	.20000		4.0000

Sample Name: ICSAB Acquired: 6/24/2016 0:04:29 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	kF **.04829**	k **-.00766**	kF **.05850**	kF **.04638**	kF **.02738**	k **-.02638**
Stddev	.00365	.00766	.00306	.00425	.00303	.00499
%RSD	7.5565	99.904	5.2255	9.1610	11.057	18.931
#1	k .04412	k -.00318	k .05693	k .04159	k .03052	k -.02995
#2	k .05085	k -.00330	k .05656	k .04971	k .02448	k -.02852
#3	k .04991	k -.01650	k .06203	k .04783	k .02714	k -.02067
Check ?	Chk Fail	Chk Pass	Chk Fail	Chk Fail	Chk Fail	Chk Pass
High Limit	.60000		.60000	.60000	.30000	
Low Limit	.40000		.40000	.40000	.20000	

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	k **-.00052**	F **.44241**	kF **29.066**	kF **.04366**	sF **.01717**	kF **.04537**
Stddev	.00035	.38366	66.439	.00317	.02029	.00340
%RSD	66.984	86.721	228.58	7.2485	118.19	7.4898
#1	k -.00035	.66466	k 105.09	k .04023	s -.00018	k .04144
#2	k -.00091	.66318	k -17.882	k .04647	s .01220	k .04733
#3	k -.00029	-.00061	k -.0063	k .04429	s .03948	k .04732
Check ?	Chk Pass	Chk Fail	Chk Fail	Chk Fail	Chk Fail	Chk Fail
High Limit		.01000	.03000	.60000	.30000	.60000
Low Limit		-.01000	-.03000	.40000	.20000	.40000

Elem	Zr3391
Units	ppm
Avg	kF **4441.9**
Stddev	9705.7
%RSD	218.50
#1	k 15574.
#2	k -2248.1
#3	k .15
Check ?	Chk Fail
High Limit	.02500
Low Limit	-.02500

Sample Name: ICSAB Acquired: 6/24/2016 0:04:29 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9862.8	^ *****	713.41
Stddev	938.9	-----	1233.6
%RSD	9.5192	-----	172.91
#1	10665.	^ -----	-.471
#2	10093.	^ -----	2.867
#3	8830.2	^ -----	2137.8

Sample Name: CCV Acquired: 6/24/2016 0:08:22 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40571**	**10.252**	**.40176**	**.50352**	**1.0259**	**.05216**	**10.249**
Stddev	.00299	.036	.00436	.00116	.0046	.00015	.046
%RSD	.73722	.35335	1.0850	.22983	.44349	.28067	.44972
#1	.40902	10.294	.39676	.50219	1.0215	.05229	10.217
#2	.40321	10.228	.40475	.50401	1.0306	.05200	10.229
#3	.40489	10.234	.40377	.50434	1.0255	.05219	10.302
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05061**	**.20363**	**.51898**	**.51228**	**4.1370**	**50.997**	**1.0462**
Stddev	.00046	.00031	.00258	.00236	.0140	.262	.0028
%RSD	.90962	.15400	.49724	.46108	.33750	.51330	.27175
#1	.05077	.20360	.52152	.50960	4.1210	50.850	1.0429
#2	.05096	.20396	.51636	.51314	4.1438	51.299	1.0477
#3	.05008	.20334	.51907	.51408	4.1463	50.841	1.0479
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**10.348**	**.51498**	**1.0223**	**50.946**	**.51068**	**10.026**	**.51362**
Stddev	.113	.00344	.0028	.211	.00210	.018	.00135
%RSD	1.0914	.66705	.27479	.41321	.41141	.17759	.26270
#1	10.350	.51110	1.0252	50.805	.51092	10.041	.51220
#2	10.460	.51621	1.0220	51.188	.50847	10.031	.51488
#3	10.234	.51763	1.0196	50.845	.51265	10.006	.51378
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Sample Name: CCV Acquired: 6/24/2016 0:08:22 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2074**	**.40666**	**5.0928**	**1.0152**	**1.0270**	**1.0159**	**.51326**
Stddev	.0053	.00225	.0089	.0020	.0057	.0060	.00130
%RSD	.43925	.55339	.17475	.20078	.55230	.59351	.25324
#1	1.2134	.40912	5.0829	1.0130	1.0205	1.0208	.51297
#2	1.2053	.40614	5.1001	1.0158	1.0309	1.0178	.51469
#3	1.2034	.40471	5.0956	1.0169	1.0297	1.0092	.51214
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0354**	**1.0194**	F **.15885**
Stddev	.0015	.0011	.28455
%RSD	.14387	.10553	179.12
#1	1.0371	1.0182	.48730
#2	1.0345	1.0199	-.01296
#3	1.0347	1.0202	.00222
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			-10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6365.1**	**39110.**	**2005.2**
Stddev	24.8	486.	19.1
%RSD	.39033	1.2423	.95151
#1	6367.8	38552.	2019.3
#2	6388.5	39438.	1983.5
#3	6339.0	39340.	2012.9

Sample Name: CCB Acquired: 6/24/2016 0:12:00 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00050**	**.01206**	**.00160**	**.00292**	**-.00076**	**.00002**	**.02662**	**.00027**
Stddev	.00084	.00666	.00259	.00138	.00101	.00006	.04195	.00026
%RSD	168.40	55.161	162.13	47.221	133.70	271.30	157.60	96.736
#1	-.00086	.01566	-.00084	.00208	-.00074	.00001	-.00269	.00055
#2	-.00111	.01615	.00432	.00217	.00024	.00009	.07467	.00003
#3	.00046	.00439	.00131	.00451	-.00178	-.00003	.00787	.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00031**	**-.00012**	**.00191**	**.00930**	**-.00712**	**-.00214**	**.08690**	**.00263**
Stddev	.00007	.00094	.00113	.01540	.09612	.00388	.02902	.00322
%RSD	21.273	789.62	59.288	165.71	1349.4	181.57	33.396	122.26
#1	.00025	.00095	.00248	.00841	-.09353	-.00318	.11978	.00265
#2	.00039	-.00048	.00263	-.00565	-.02424	.00216	.06485	.00584
#3	.00031	-.00083	.00060	.02512	.09640	-.00539	.07607	-.00060
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00355**	**.02859**	**-.00197**	**.00635**	**-.00184**	**.00648**	**-.00780**	**.00562**
Stddev	.00019	.02702	.00094	.00761	.00466	.00302	.00403	.00110
%RSD	5.3403	94.506	47.726	119.74	253.74	46.674	51.730	19.667
#1	.00349	.02662	-.00294	-.00235	-.00211	.00986	-.01231	.00668
#2	.00339	.00261	-.00192	.01175	-.00636	.00555	-.00455	.00447
#3	.00376	.05655	-.00106	.00967	.00296	.00403	-.00652	.00571
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: CCB Acquired: 6/24/2016 0:12:00 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00027	.00037	-.00205	-.00398	-.00052	.00005	-.03486
Stddev	.00199	.00049	.00712	.00235	.00010	.00013	.29949
%RSD	737.85	134.70	347.62	58.933	18.924	280.33	859.07
#1	.00244	-.00019	-.00912	-.00572	-.00063	.00006	.11572
#2	-.00149	.00054	.00512	-.00493	-.00049	-.00009	.15945
#3	-.00013	.00074	-.00215	-.00131	-.00044	.00017	-.37976
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	6385.3	39685.	2013.6
Stddev	13.4	625.	15.5
%RSD	.20981	1.5744	.76988
#1	6370.9	40406.	2004.4
#2	6397.5	39350.	2031.5
#3	6387.5	39298.	2004.9

Sample Name: LLCCV Acquired: 6/24/2016 0:16:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00692	.17280	.00777	.08023	.00816	.00159	.47649	.00097
Stddev	.00020	.00275	.00190	.00358	.00068	.00009	.02580	.00026
%RSD	2.8583	1.5889	24.477	4.4629	8.3509	5.9617	5.4139	26.520
#1	.00710	.16995	.00825	.08038	.00802	.00170	.44675	.00126
#2	.00671	.17543	.00940	.08374	.00756	.00154	.48987	.00076
#3	.00694	.17303	.00568	.07658	.00890	.00153	.49285	.00090
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00419	.00395	.00595	.08433	.74272	.08170	.43560	.00898
Stddev	.00055	.00077	.00059	.03278	.15549	.00079	.02792	.00231
%RSD	13.081	19.592	9.9605	38.876	20.936	.96348	6.4091	25.784
#1	.00481	.00337	.00657	.12211	.85231	.08123	.41810	.00714
#2	.00397	.00365	.00539	.06743	.81110	.08127	.46780	.01158
#3	.00379	.00483	.00590	.06344	.56476	.08261	.42090	.00822
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00841	.43808	.01603	.80173	.00604	.08746	.00868	.86554
Stddev	.00048	.03267	.00038	.00808	.00286	.00565	.00374	.00329
%RSD	5.6889	7.4567	2.3907	1.0073	47.406	6.4564	43.147	.37989
#1	.00807	.47434	.01600	.80832	.00935	.08147	.01238	.86709
#2	.00820	.41094	.01567	.80416	.00429	.08820	.00489	.86176
#3	.00895	.42897	.01643	.79272	.00448	.09269	.00877	.86777
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Sample Name: LLCCV Acquired: 6/24/2016 0:16:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41147**	**.04220**	**.01937**	**.15204**	**.00784**	**.01708**	**20.869**
Stddev	.00081	.00045	.00562	.00340	.00117	.00020	.318
%RSD	.19753	1.0715	29.021	2.2390	14.946	1.1523	1.5231
#1	.41153	.04213	.02561	.15455	.00880	.01720	21.227
#2	.41224	.04269	.01469	.15340	.00817	.01718	20.618
#3	.41062	.04179	.01783	.14816	.00653	.01685	20.762
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6476.0**	**40121.**	**2010.8**
Stddev	41.9	341.	10.1
%RSD	.64677	.85050	.50105
#1	6439.6	40107.	2010.7
#2	6521.8	39788.	2020.9
#3	6466.6	40470.	2000.7

Sample Name: LLCCV Acquired: 6/24/2016 0:19:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02136	.41629	.02112	.19510	.02080	.00407	1.0587
Stddev	.00041	.00284	.00410	.00052	.00091	.00010	.0154
%RSD	1.9202	.68293	19.397	.26774	4.3671	2.4437	1.4499
#1	.02183	.41875	.01725	.19543	.02070	.00418	1.0490
#2	.02111	.41695	.02541	.19450	.02176	.00402	1.0764
#3	.02114	.41318	.02071	.19539	.01995	.00401	1.0506
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00191	.01019	.00970	.01209	.22512	2.0312	.20560
Stddev	.00008	.00029	.00094	.00032	.02751	.1478	.00292
%RSD	3.9464	2.8921	9.6498	2.6079	12.219	7.2758	1.4184
#1	.00189	.00987	.00862	.01175	.23193	1.9599	.20462
#2	.00184	.01046	.01027	.01237	.19484	2.2011	.20331
#3	.00199	.01022	.01020	.01214	.24858	1.9325	.20889
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.0419	.02143	.01974	1.0634	.04092	1.9580	.02354
Stddev	.1407	.00617	.00054	.0338	.00119	.0082	.00145
%RSD	13.508	28.783	2.7554	3.1802	2.8968	.41842	6.1627
#1	1.1704	.01821	.01917	1.0842	.04126	1.9586	.02462
#2	1.0638	.01754	.01982	1.0815	.03960	1.9658	.02189
#3	.89149	.02854	.02024	1.0243	.04189	1.9495	.02411
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Sample Name: LLCCV Acquired: 6/24/2016 0:19:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v969) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.21053**	**.03539**	**2.1191**	**1.0135**	**.10298**	**.05480**	**.37888**
Stddev	.00225	.00798	.0028	.0011	.00044	.01048	.00446
%RSD	1.0680	22.551	.13372	.10864	.43012	19.119	1.1784
#1	.21009	.02618	2.1179	1.0144	.10249	.04402	.38327
#2	.21296	.04014	2.1223	1.0138	.10312	.05543	.37434
#3	.20853	.03987	2.1170	1.0123	.10334	.06494	.37903
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.01937**	**.04203**	F **55.206**
Stddev	.00081	.00024	.600
%RSD	4.1646	.56559	1.0870
#1	.02030	.04231	55.899
#2	.01893	.04187	54.851
#3	.01888	.04193	54.868
Check ?	Chk Pass	Chk Pass	Chk Fail
High Limit			36.000
Low Limit			-.04000

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**6500.2**	**40568.**	**2033.1**
Stddev	36.1	206.	33.6
%RSD	.55595	.50832	1.6527
#1	6459.1	40580.	1998.8
#2	6514.4	40767.	2065.9
#3	6527.1	40355.	2034.6

Element, Wavelength and Order	Date of Fit	Date of Cal.	Type of Fit	Weighting	A0	A1	A2	n (Exponent)
Ag 328.068 {103}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000236	0.051213	0.000000	1.000000
Al 308.215 {109}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000796	0.008044	0.000000	1.000000
As 189.042 {478}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000007	0.025071	0.000000	1.000000
B 249.678 {135}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000040	0.023963	0.000000	1.000000
Ba 455.403 { 74}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.044484	2.168568	0.000000	1.000000
Be 313.107 {108}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000479	0.930119	0.000000	1.000000
Ca 422.673 { 80}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.002874	0.045862	0.000000	1.000000
Cd 228.802 {447}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000038	0.570455	0.000000	1.000000
Co 228.616 {447}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000129	0.339574	0.000000	1.000000
Cr 267.716 {126}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000254	0.058743	0.000000	1.000000
Cu 224.700 {450}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000063	0.129645	0.000000	1.000000
Fe 261.187 {129}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000714	0.024119	0.000000	1.000000
K 766.490 { 44}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.006604	0.061274	0.000000	1.000000
Li 670.784 { 50}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.009415	1.520559	0.000000	1.000000
Mg 279.079 {121}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000260	0.008000	0.000000	1.000000
Mn 257.610 {131}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000470	0.289814	0.000000	1.000000
Mo 202.030 {467}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000054	0.182736	0.000000	1.000000
Na 589.592 { 57}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.047777	0.184477	0.000000	1.000000
Ni 231.604 {446}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000734	0.138990	0.000000	1.000000
P 214.914 {457}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000172	0.014319	0.000000	1.000000
Pb 220.353 {453}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.001186	0.065283	0.000000	1.000000
Sb 206.833 {463}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000368	0.031084	0.000000	1.000000
Se 196.090 {472}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000508	0.017412	0.000000	1.000000
Si 212.412 {459}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.001467	0.039573	0.000000	1.000000
Sn 189.989 {477}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000022	0.066307	0.000000	1.000000
Sr 407.771 { 83}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.001060	3.591964	0.000000	1.000000
Ti 337.280 {100}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.002630	0.122957	0.000000	1.000000
Tl 190.856 {477}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000329	0.028655	0.000000	1.000000
V 292.402 {115}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.000071	0.067328	0.000000	1.000000
Y 224.306 {450}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Y 360.073 { 94}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Y 377.433 { 89}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Zn 206.200 {463}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	0.000142	0.437417	0.000000	1.000000
Zr 339.198 { 99}	6/25/2016 13:11:09	6/25/2016 12:40:05	Linear	1/Conc	-0.007106	1.219765	0.000000	1.000000

Approved: June 27, 2016

Element, Wavelength and Order	Correlation	Std Error of Est	Predicted MDL	Predicted MQL	Status	Reslope		QC Norm	
						Slope	Y-int	Slope factor	Offset
Ag 328.068 {103}	0.999680	0.000003	0.002344	0.007814	OK.	1.000000	0.000000	1	0
Al 308.215 {109}	0.999950	0.000005	0.010188	0.033959	OK.	1.000000	0.000000	1	0
As 189.042 {478}	0.999909	0.000001	0.003189	0.010630	OK.	1.000000	0.000000	1	0
B 249.678 {135}	0.999961	0.000001	0.002963	0.009878	OK.	1.000000	0.000000	1	0
Ba 455.403 { 74}	0.999951	0.000136	0.001119	0.003729	OK.	1.000000	0.000000	1	0
Be 313.107 {108}	0.999766	0.000007	0.000105	0.000348	OK.	1.000000	0.000000	1	0
Ca 422.673 { 80}	0.999890	0.000043	0.043792	0.145974	OK.	1.000000	0.000000	1	0
Cd 228.802 {447}	0.999970	0.000002	0.000282	0.000940	OK.	1.000000	0.000000	1	0
Co 228.616 {447}	0.999977	0.000003	0.000454	0.001514	OK.	1.000000	0.000000	1	0
Cr 267.716 {126}	0.999978	0.000001	0.001307	0.004355	OK.	1.000000	0.000000	1	0
Cu 224.700 {450}	0.999889	0.000006	0.001405	0.004682	OK.	1.000000	0.000000	1	0
Fe 261.187 {129}	0.999330	0.000022	0.023003	0.076675	OK.	1.000000	0.000000	1	0
K 766.490 { 44}	0.999988	0.000096	0.104199	0.347330	OK.	1.000000	0.000000	1	0
Li 670.784 { 50}	0.999977	0.000100	0.004201	0.014002	OK.	1.000000	0.000000	1	0
Mg 279.079 {121}	0.999990	0.000003	0.092396	0.307987	OK.	1.000000	0.000000	1	0
Mn 257.610 {131}	0.999236	0.000036	0.002652	0.008839	OK.	1.000000	0.000000	1	0
Mo 202.030 {467}	0.999956	0.000011	0.000450	0.001500	OK.	1.000000	0.000000	1	0
Na 589.592 { 57}	0.999990	0.000264	0.032289	0.107629	OK.	1.000000	0.000000	1	0
Ni 231.604 {446}	0.999932	0.000005	0.001180	0.003935	OK.	1.000000	0.000000	1	0
P 214.914 {457}	0.999990	0.000004	0.007138	0.023792	OK.	1.000000	0.000000	1	0
Pb 220.353 {453}	0.999383	0.000007	0.003667	0.012223	OK.	1.000000	0.000000	1	0
Sb 206.833 {463}	0.999854	0.000004	0.004516	0.015054	OK.	1.000000	0.000000	1	0
Se 196.090 {472}	0.999238	0.000003	0.006704	0.022346	OK.	1.000000	0.000000	1	0
Si 212.412 {459}	0.999594	0.000036	0.002443	0.008145	OK.	1.000000	0.000000	1	0
Sn 189.989 {477}	0.999980	0.000003	0.000986	0.003287	OK.	1.000000	0.000000	1	0
Sr 407.771 { 83}	0.999952	0.000224	0.000467	0.001557	OK.	1.000000	0.000000	1	0
Ti 337.280 {100}	0.999579	0.000023	0.007453	0.024845	OK.	1.000000	0.000000	1	0
Tl 190.856 {477}	0.999831	0.000002	0.003284	0.010945	OK.	1.000000	0.000000	1	0
V 292.402 {115}	0.999969	0.000003	0.001095	0.003650	OK.	1.000000	0.000000	1	0
Y 224.306 {450}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Y 360.073 { 94}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Y 377.433 { 89}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Zn 206.200 {463}	0.999941	0.000030	0.000202	0.000675	OK.	1.000000	0.000000	1	0
Zr 339.198 { 99}	0.596446	0.010414	0.001594	0.005312	OK.	1.000000	0.000000	1	0

Approved: June 27, 2016

Sample Name: S0 Acquired: 6/25/2016 12:21:04 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00024**	**.00080**	**-.00001**	**.00004**	**.04448**	**.00048**	**.00288**
Stddev	.00008	.00004	.00007	.00002	.00170	.00005	.00169
%RSD	35.712	4.5527	977.99	54.959	3.8148	10.198	58.629
#1	-.00024	.00076	-.00006	.00003	.04635	.00042	.00096
#2	-.00015	.00081	.00007	.00006	.04405	.00050	.00353
#3	-.00032	.00083	-.00004	.00003	.04304	.00052	.00414

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00004**	**-.00013**	**.00025**	**-.00006**	**-.00071**	**-.00661**	**-.00941**
Stddev	.00006	.00005	.00003	.00023	.00026	.00355	.00965
%RSD	153.97	36.064	10.924	367.85	35.978	53.724	102.48
#1	-.00002	-.00016	.00028	-.00022	-.00093	-.00948	-.01508
#2	-.00010	-.00015	.00025	.00020	-.00043	-.00769	.00173
#3	.00001	-.00008	.00023	-.00018	-.00078	-.00264	-.01488

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00026**	**.00047**	**.00005**	**-.04778**	**-.00073**	**-.00017**	**-.00119**
Stddev	.00066	.00061	.00004	.00421	.00008	.00002	.00008
%RSD	255.00	128.77	78.942	8.8105	11.011	13.180	6.4052
#1	.00045	-.00015	.00007	-.05263	-.00065	-.00015	-.00110
#2	-.00086	.00049	.00001	-.04509	-.00082	-.00019	-.00123
#3	-.00037	.00107	.00009	-.04563	-.00073	-.00017	-.00123

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00037**	**-.00051**	**.00147**	**.00002**	**.00106**	**-.00263**	**-.00033**
Stddev	.00001	.00007	.00009	.00003	.00137	.00025	.00002
%RSD	3.7153	13.815	6.2921	126.34	130.24	9.4020	7.1687
#1	.00035	-.00043	.00157	.00002	.00203	-.00235	-.00030
#2	.00038	-.00057	.00145	-.00000	.00165	-.00283	-.00033
#3	.00037	-.00052	.00138	.00005	-.00052	-.00270	-.00035

Approved: June 27, 2016

Sample Name: S0 Acquired: 6/25/2016 12:21:04 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S
Avg	-.00007	.00014	-.00710
Stddev	.00004	.00005	.00324
%RSD	58.279	35.146	45.700
#1	-.00004	.00009	-.00457
#2	-.00005	.00016	-.01076
#3	-.00012	.00018	-.00597
Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9736.8	58329.	2962.8
Stddev	24.6	275.	12.4
%RSD	.25315	.47180	.41783
#1	9719.4	58177.	2949.0
#2	9725.9	58162.	2966.1
#3	9765.0	58646.	2973.1

Approved: June 27, 2016

Sample Name: S1 Acquired: 6/25/2016 12:25:02 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	Ba4554	Be3131	Ca4226	Cd2288	Co2286
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00000**	**.00148**	**.06336**	**.00097**	**.00582**	**.00018**	**.00038**
Stddev	.00004	.00005	.00419	.00009	.00234	.00011	.00006
%RSD	838.91	3.6655	6.6080	8.8478	40.284	60.110	16.615
#1	.00001	.00145	.05987	.00105	.00569	.00026	.00043
#2	-.00004	.00145	.06800	.00097	.00822	.00006	.00031
#3	.00004	.00155	.06221	.00088	.00354	.00023	.00041

Elem	Cr2677	Cu2247	Fe2611	K_7664	Mn2576	Mo2020	Na5895
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00050**	**.00053**	**.00033**	**.01950**	**.00077**	**.00143**	**.02669**
Stddev	.00003	.00011	.00021	.00355	.00009	.00011	.00332
%RSD	5.5692	20.305	62.737	18.223	11.862	7.4791	12.449
#1	.00049	.00060	.00053	.01550	.00083	.00145	.02356
#2	.00048	.00041	.00012	.02070	.00066	.00152	.03018
#3	.00053	.00058	.00034	.02230	.00081	.00131	.02634

Elem	Ni2316	P_2149	Pb2203	Sb2068	Si2124	Sn1899	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00018**	**.00100**	**-.00086**	**.00059**	**.00297**	**.00055**	**.03302**
Stddev	.00009	.00012	.00030	.00014	.00005	.00006	.00089
%RSD	52.159	12.366	34.406	23.539	1.7045	10.953	2.6942
#1	-.00010	.00113	-.00102	.00051	.00294	.00050	.03225
#2	-.00028	.00098	-.00052	.00075	.00302	.00062	.03399
#3	-.00016	.00088	-.00105	.00050	.00293	.00055	.03280

Elem	Ti3372	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00218**	**.00050**	**.00363**	**-.00591**
Stddev	.00071	.00003	.00003	.00134
%RSD	32.781	5.1557	.77751	22.659
#1	-.00139	.00050	.00365	-.00576
#2	-.00277	.00048	.00360	-.00465
#3	-.00237	.00053	.00364	-.00732

Approved: June 27, 2016

Sample Name: S1 Acquired: 6/25/2016 12:25:02 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9839.2**	**59021.**	**2974.6**
Stddev	35.8	140.	1.9
%RSD	.36397	.23691	.06440
#1	9811.4	58880.	2976.5
#2	9826.5	59023.	2972.7
#3	9879.6	59160.	2974.8

Approved: June 27, 2016

Sample Name: S2 Acquired: 6/25/2016 12:29:00 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00011**	**.00220**	**.00018**	**.00025**	**.08112**	**.00139**	**.00931**
Stddev	.00003	.00006	.00005	.00002	.00125	.00003	.00047
%RSD	29.658	2.5782	25.201	8.0082	1.5417	1.8740	5.0396
#1	.00012	.00221	.00013	.00026	.08191	.00140	.00905
#2	.00012	.00214	.00022	.00027	.08177	.00136	.00985
#3	.00007	.00225	.00019	.00023	.07968	.00140	.00903

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00041**	**.00102**	**.00072**	**.00104**	**.00148**	**.04190**	**.01415**
Stddev	.00007	.00003	.00001	.00014	.00030	.00236	.00012
%RSD	16.350	3.0292	1.5873	13.319	20.240	5.6312	.86545
#1	.00045	.00101	.00072	.00109	.00137	.04369	.01405
#2	.00033	.00099	.00073	.00088	.00125	.03923	.01429
#3	.00045	.00105	.00071	.00115	.00182	.04279	.01411

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00109**	**.00274**	**.00298**	**.10797**	**.00048**	**.00211**	**-.00045**
Stddev	.00041	.00057	.00005	.00406	.00013	.00008	.00017
%RSD	37.607	20.931	1.7509	3.7561	26.749	3.7519	39.075
#1	.00144	.00213	.00304	.11260	.00057	.00211	-.00064
#2	.00064	.00283	.00295	.10501	.00033	.00203	-.00031
#3	.00120	.00327	.00295	.10631	.00054	.00219	-.00039

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00105**	**-.00034**	**.00448**	**.00108**	**.06082**	**-.00047**	**-.00014**
Stddev	.00006	.00004	.00009	.00004	.00131	.00089	.00005
%RSD	6.0784	12.562	2.0565	3.6593	2.1502	187.41	36.276
#1	.00109	-.00029	.00447	.00110	.05950	-.00104	-.00016
#2	.00107	-.00037	.00458	.00110	.06212	-.00093	-.00008
#3	.00097	-.00035	.00440	.00103	.06083	.00055	-.00017

Approved: June 27, 2016

Sample Name: S2 Acquired: 6/25/2016 12:29:00 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S
Avg	.00105	.00728	-.00437
Stddev	.00006	.00006	.00132
%RSD	5.7588	.85058	30.251
#1	.00102	.00735	-.00588
#2	.00112	.00727	-.00376
#3	.00101	.00722	-.00345
Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9827.0	59373.	3004.6
Stddev	11.4	114.	4.0
%RSD	.11574	.19273	.13472
#1	9814.0	59255.	3000.6
#2	9832.7	59484.	3008.7
#3	9834.5	59381.	3004.6

Approved: June 27, 2016

Sample Name: S3 Acquired: 6/25/2016 12:32:56 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.02035	.08278	.00992	.01198	2.2359	.04820	.46599	.03126
Stddev	.00007	.00029	.00007	.00002	.0069	.00016	.00236	.00006
%RSD	.36344	.34990	.73624	.13878	.30693	.33862	.50730	.19965
#1	.02029	.08301	.00987	.01198	2.2433	.04825	.46861	.03126
#2	.02044	.08286	.01000	.01199	2.2346	.04833	.46401	.03132
#3	.02033	.08245	.00988	.01196	2.2297	.04802	.46536	.03119

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.06810	.02991	.06581	.09648	3.0722	1.5250	.07962	.14725
Stddev	.00012	.00021	.00002	.00063	.0027	.0048	.00098	.00051
%RSD	.17879	.70022	.03611	.65013	.08880	.31624	1.2353	.34909
#1	.06804	.03009	.06584	.09719	3.0745	1.5268	.07965	.14783
#2	.06801	.02997	.06579	.09600	3.0692	1.5288	.07861	.14685
#3	.06824	.02968	.06581	.09626	3.0730	1.5196	.08058	.14706

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.18510	9.1987	.06966	.14157	.03222	.03693	.00649	.20808
Stddev	.00068	.0276	.00008	.00029	.00023	.00011	.00003	.00004
%RSD	.36712	.29975	.11504	.20555	.70478	.30304	.43769	.01830
#1	.18567	9.2297	.06963	.14170	.03236	.03694	.00648	.20811
#2	.18529	9.1894	.06975	.14177	.03196	.03703	.00652	.20808
#3	.18435	9.1769	.06959	.14123	.03234	.03681	.00647	.20804

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.06680	3.6291	.12115	.01313	.06767	.44396	3.5474
Stddev	.00012	.0161	.00039	.00012	.00004	.00054	.0233
%RSD	.17282	.44347	.32200	.91385	.05243	.12075	.65596
#1	.06677	3.6476	.12157	.01325	.06770	.44451	3.5656
#2	.06671	3.6220	.12107	.01301	.06768	.44393	3.5554
#3	.06693	3.6179	.12080	.01311	.06763	.44344	3.5212

Approved: June 27, 2016

Sample Name: S3 Acquired: 6/25/2016 12:32:56 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9929.1	59890.	3083.9
Stddev	18.0	90.	16.7
%RSD	.18173	.14986	.54188
#1	9944.3	59787.	3069.2
#2	9933.7	59947.	3080.5
#3	9909.1	59936.	3102.1

Approved: June 27, 2016

Sample Name: S4 Acquired: 6/25/2016 12:36:34 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.04037	.16212	.01994	.02391	4.3553	.09476	.91725	.06225
Stddev	.00010	.00009	.00010	.00008	.0282	.00010	.00714	.00008
%RSD	.26004	.05310	.51424	.32854	.64632	.10278	.77855	.12431
#1	.04040	.16220	.02005	.02399	4.3445	.09472	.91430	.06220
#2	.04046	.16203	.01991	.02383	4.3342	.09469	.91205	.06220
#3	.04026	.16211	.01985	.02390	4.3873	.09487	.92539	.06233

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.13502	.05877	.12834	.19056	6.1046	3.0186	.15914	.28946
Stddev	.00023	.00007	.00015	.00161	.0456	.0309	.00223	.00202
%RSD	.17276	.12518	.11394	.84544	.74756	1.0253	1.3995	.69732
#1	.13475	.05870	.12846	.19050	6.0869	3.0140	.15736	.28818
#2	.13515	.05878	.12838	.18897	6.0705	2.9902	.15842	.28841
#3	.13517	.05884	.12818	.19219	6.1564	3.0516	.16164	.29179

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.36324	18.368	.13725	.28094	.06365	.07367	.01336	.39092
Stddev	.00176	.146	.00020	.00054	.00019	.00012	.00006	.00048
%RSD	.48547	.79728	.14492	.19364	.30628	.15884	.42128	.12331
#1	.36486	18.303	.13728	.28153	.06355	.07380	.01339	.39147
#2	.36349	18.266	.13744	.28083	.06388	.07365	.01329	.39059
#3	.36136	18.536	.13705	.28046	.06354	.07357	.01339	.39071

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.13179	7.1439	.24230	.02560	.13357	.86773	.12255
Stddev	.00029	.0568	.00280	.00005	.00025	.00172	.00222
%RSD	.21785	.79476	1.1551	.20145	.18401	.19856	1.8114
#1	.13210	7.1181	.24103	.02555	.13357	.86947	.12020
#2	.13172	7.1045	.24036	.02563	.13332	.86769	.12285
#3	.13154	7.2090	.24551	.02564	.13382	.86603	.12461

Approved: June 27, 2016

Sample Name: S4 Acquired: 6/25/2016 12:36:34 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9795.6**	**59125.**	**3064.9**
Stddev	6.6	26.	19.7
%RSD	.06757	.04412	.64222
#1	9799.1	59120.	3071.1
#2	9799.7	59153.	3080.8
#3	9787.9	59102.	3042.9

Approved: June 27, 2016

Sample Name: ICV Acquired: 6/25/2016 12:40:09 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.40482	10.096	.40189	.50535	1.0085	.05038	10.096
Stddev	.00169	.009	.00320	.00259	.0096	.00007	.103
%RSD	.41792	.09100	.79651	.51179	.95284	.13225	1.0175
#1	.40318	10.095	.40558	.50259	.99972	.05037	10.065
#2	.40656	10.106	.40020	.50771	1.0070	.05032	10.013
#3	.40473	10.088	.39988	.50575	1.0188	.05045	10.211
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05022	.20172	.50610	.50772	4.0459	50.287	1.0069
Stddev	.00031	.00057	.00078	.00110	.0398	.526	.0059
%RSD	.61395	.28458	.15454	.21701	.98466	1.0464	.58082
#1	.05048	.20224	.50644	.50864	4.0020	49.775	1.0012
#2	.05030	.20181	.50521	.50650	4.0559	50.260	1.0067
#3	.04988	.20110	.50666	.50803	4.0798	50.827	1.0128
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	10.117	.50795	1.0200	50.068	.50832	10.053	.50460
Stddev	.152	.00312	.0050	.377	.00054	.029	.00016
%RSD	1.5029	.61369	.49437	.75311	.10596	.28628	.03077
#1	9.9425	.50440	1.0241	49.665	.50829	10.073	.50452
#2	10.219	.50922	1.0214	50.126	.50888	10.066	.50449
#3	10.191	.51024	1.0144	50.412	.50780	10.020	.50477
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: ICV Acquired: 6/25/2016 12:40:09 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2132**	**.39981**	**5.2115**	**1.0100**	**1.0094**	**1.0076**	**.50967**
Stddev	.0090	.00367	.0220	.0047	.0093	.0117	.00340
%RSD	.74499	.91900	.42193	.46646	.92212	1.1612	.66719
#1	1.2228	.39652	5.1880	1.0139	.99976	.99418	.50608
#2	1.2117	.40378	5.2316	1.0112	1.0100	1.0128	.51284
#3	1.2049	.39915	5.2149	1.0048	1.0183	1.0157	.51010
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0117**	**1.0154**	F **2.9010**
Stddev	.0015	.0015	.0057
%RSD	.14666	.14385	.19619
#1	1.0111	1.0164	2.9026
#2	1.0134	1.0160	2.8946
#3	1.0106	1.0137	2.9057
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			5.0000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9887.2**	**59604.**	**3090.1**
Stddev	2.6	87.	8.3
%RSD	.02633	.14631	.26858
#1	9888.5	59705.	3098.3
#2	9884.2	59557.	3090.3
#3	9888.9	59551.	3081.7

Approved: June 27, 2016

Sample Name: ICB Acquired: 6/25/2016 12:43:48 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00131	-.00019	.00035	.00523	.00097	.00005	-.00769	-.00024
Stddev	.00204	.00299	.00150	.00175	.00089	.00006	.00734	.00011
%RSD	155.46	1561.3	426.44	33.377	91.849	115.36	95.505	46.996
#1	.00317	-.00362	.00181	.00349	.00060	.00012	-.01556	-.00036
#2	-.00088	.00114	-.00118	.00522	.00198	.00002	-.00104	-.00017
#3	.00165	.00191	.00042	.00699	.00032	.00001	-.00646	-.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00000	.00019	.00069	.02190	.04427	.00252	.00787	.00050
Stddev	.00033	.00096	.00130	.00895	.02254	.00192	.08048	.00146
%RSD	40694.	494.95	189.51	40.889	50.920	76.061	1022.8	292.29
#1	-.00026	-.00067	-.00034	.02348	.04464	.00424	-.08504	-.00108
#2	-.00012	.00002	.00215	.01226	.06662	.00045	.05249	.00079
#3	.00038	.00123	.00025	.02996	.02154	.00288	.05615	.00179
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00437	.07087	.00070	.00285	.00118	.00678	.00870	.00121
Stddev	.00041	.01151	.00054	.00868	.00197	.00381	.00450	.00189
%RSD	9.2912	16.248	77.044	304.78	167.53	56.206	51.723	156.22
#1	.00395	.08310	.00020	-.00283	.00108	.00337	.00854	.00310
#2	.00476	.06927	.00064	-.00147	.00320	.00608	.00428	.00120
#3	.00442	.06023	.00127	.01284	-.00074	.01090	.01328	-.00068
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: ICB Acquired: 6/25/2016 12:43:48 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00063**	**.00059**	**.00116**	**.00143**	**.00111**	**.00117**	**.00345**
Stddev	.00049	.00006	.00559	.00387	.00033	.00016	.00039
%RSD	77.601	9.5579	481.47	269.75	29.521	13.916	11.209
#1	.00050	.00059	.00521	.00260	.00135	.00105	.00361
#2	.00117	.00064	.00349	-.00288	.00074	.00111	.00300
#3	.00022	.00053	-.00522	.00459	.00124	.00135	.00372
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9709.0**	**58834.**	**2997.3**
Stddev	13.6	105.	11.0
%RSD	.13963	.17854	.36570
#1	9693.7	58895.	2985.3
#2	9713.8	58895.	3006.9
#3	9719.5	58713.	2999.6

Approved: June 27, 2016

Sample Name: LLICV Acquired: 6/25/2016 12:47:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00947**	**.16517**	**.00750**	**.08209**	**.00794**	**.00166**	**.41368**	**.00078**
Stddev	.00234	.00914	.00427	.00134	.00096	.00002	.01542	.00015
%RSD	24.742	5.5361	56.913	1.6344	12.112	1.4530	3.7271	18.963
#1	.01038	.17315	.00639	.08214	.00810	.00169	.40911	.00061
#2	.01123	.15519	.01222	.08341	.00690	.00166	.40107	.00083
#3	.00681	.16717	.00390	.08073	.00880	.00164	.43087	.00089
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00435**	**.00476**	**.00462**	**.09393**	**.88733**	**.08419**	**.44121**	**.00701**
Stddev	.00017	.00091	.00136	.00523	.14622	.00363	.12107	.00237
%RSD	3.8803	19.165	29.360	5.5689	16.479	4.3143	27.441	33.765
#1	.00455	.00411	.00608	.08852	.75945	.08105	.39471	.00889
#2	.00425	.00438	.00436	.09433	.85580	.08335	.57864	.00779
#3	.00425	.00581	.00341	.09896	1.0467	.08816	.35028	.00435
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00866**	**.44156**	**.01715**	**.79893**	**.00705**	**.08442**	**.02038**	**.79310**
Stddev	.00052	.01049	.00062	.00929	.00266	.00215	.00191	.00106
%RSD	5.9984	2.3750	3.6174	1.1626	37.766	2.5524	9.3696	.13338
#1	.00809	.45038	.01661	.80194	.00796	.08678	.01818	.79326
#2	.00879	.42997	.01783	.80634	.00405	.08394	.02136	.79407
#3	.00910	.44434	.01701	.78851	.00913	.08255	.02160	.79197
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LLICV Acquired: 6/25/2016 12:47:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41130**	**.04174**	**.02562**	**.15269**	**.00881**	**.01717**	**.13277**
Stddev	.00223	.00064	.00303	.00060	.00064	.00024	.00310
%RSD	.54293	1.5249	11.843	.39147	7.2679	1.3907	2.3338
#1	.41347	.04173	.02710	.15201	.00926	.01741	.13005
#2	.41142	.04111	.02213	.15292	.00808	.01715	.13211
#3	.40901	.04238	.02764	.15314	.00909	.01694	.13614
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9760.6**	**59445.**	**3032.4**
Stddev	2.7	49.	16.0
%RSD	.02790	.08308	.52659
#1	9758.1	59502.	3050.3
#2	9763.5	59418.	3027.2
#3	9760.2	59415.	3019.7

Approved: June 27, 2016

Sample Name: LLICV Acquired: 6/25/2016 12:51:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02232	.40482	.02105	.19795	.01981	.00414	.94713	.00209
Stddev	.00322	.01572	.00276	.00137	.00013	.00016	.01328	.00025
%RSD	14.431	3.8838	13.088	.69255	.67950	3.8114	1.4024	12.129
#1	.02018	.38785	.01788	.19912	.01996	.00396	.94307	.00218
#2	.02076	.40770	.02252	.19644	.01971	.00427	.93634	.00229
#3	.02603	.41890	.02277	.19828	.01975	.00419	.96196	.00181
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00992	.00950	.01082	.20111	1.9275	.20201	.97307	.01841
Stddev	.00043	.00028	.00021	.02025	.1498	.00564	.04542	.00186
%RSD	4.3042	2.9103	1.9731	10.069	7.7699	2.7925	4.6682	10.079
#1	.01005	.00935	.01089	.20826	1.8000	.20412	1.0187	.02052
#2	.00945	.00934	.01100	.21682	1.8900	.19562	.97274	.01702
#3	.01028	.00982	.01059	.17825	2.0924	.20630	.92781	.01771
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01972	1.0340	.04153	1.9482	.02271	.20123	.04411	1.9392
Stddev	.00048	.0096	.00031	.0042	.00204	.00295	.00361	.0022
%RSD	2.4591	.93077	.75288	.21380	8.9612	1.4657	8.1900	.11326
#1	.01969	1.0240	.04132	1.9520	.02496	.20436	.04094	1.9404
#2	.01925	1.0432	.04139	1.9437	.02217	.20085	.04804	1.9406
#3	.02022	1.0348	.04189	1.9488	.02100	.19850	.04333	1.9367
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LLICV Acquired: 6/25/2016 12:51:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.0099**	**.10045**	**.06134**	**.37714**	**.02003**	**.04134**	**.29456**
Stddev	.0057	.00069	.00459	.00090	.00046	.00030	.00329
%RSD	.56948	.68667	7.4856	.23976	2.2900	.72946	1.1169
#1	1.0145	.09991	.06229	.37809	.02052	.04128	.29317
#2	1.0117	.10021	.05634	.37629	.01997	.04166	.29219
#3	1.0035	.10123	.06538	.37705	.01961	.04107	.29832
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9725.0**	**59145.**	**3053.0**
Stddev	6.0	91.	15.9
%RSD	.06216	.15348	.52133
#1	9731.9	59116.	3042.4
#2	9721.3	59247.	3071.3
#3	9721.7	59072.	3045.2

Approved: June 27, 2016

Sample Name: ICSA Acquired: 6/25/2016 12:55:29 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00027**	**265.97**	**.00161**	**-.00823**	**.00056**	**.00011**	**243.96**	**-.00018**
Stddev	.00121	.27	.00172	.00296	.00012	.00009	1.89	.00042
%RSD	450.36	.10204	106.61	35.970	21.237	81.140	.77435	230.52
#1	-.00046	265.69	.00220	-.01033	.00047	.00017	241.81	-.00048
#2	.00103	265.97	-.00032	-.00484	.00069	.00014	245.35	.00030
#3	-.00137	266.24	.00295	-.00951	.00051	.00001	244.72	-.00037
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00099**	**-.00100**	**-.00098**	**96.848**	**.05329**	**.01035**	**248.71**	**-.00176**
Stddev	.00039	.00071	.00330	.692	.06120	.00255	1.99	.00096
%RSD	39.651	70.751	335.23	.71414	114.86	24.670	.80129	54.809
#1	-.00142	-.00119	.00263	96.056	.12150	.01230	246.41	-.00274
#2	-.00085	-.00160	-.00175	97.332	.03517	.00746	249.90	-.00082
#3	-.00068	-.00022	-.00383	97.155	.00318	.01129	249.83	-.00171
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00066**	**.04038**	**-.00089**	**.02257**	**.00214**	**-.00550**	**-.00224**	**.00879**
Stddev	.00023	.00410	.00050	.00580	.00444	.00318	.00753	.00258
%RSD	34.448	10.162	56.288	25.707	206.94	57.856	337.10	29.379
#1	-.00088	.04493	-.00121	.01613	.00721	-.00186	.00631	.00581
#2	-.00043	.03696	-.00115	.02418	.00024	-.00776	-.00511	.01037
#3	-.00066	.03925	-.00031	.02740	-.00103	-.00688	-.00791	.01019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: ICSA Acquired: 6/25/2016 12:55:29 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00053**	**.00100**	**-.00403**	**-.00373**	**-.00305**	**-.00596**	**-.01388**
Stddev	.00149	.00018	.00805	.00252	.00022	.00010	.00081
%RSD	284.28	18.415	199.81	67.478	7.0947	1.6059	5.8325
#1	.00144	.00117	.00514	-.00549	-.00300	-.00603	-.01379
#2	.00134	.00081	-.00728	-.00085	-.00329	-.00601	-.01311
#3	-.00120	.00101	-.00995	-.00485	-.00287	-.00585	-.01473
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9184.9**	**55126.**	**2960.9**
Stddev	19.5	129.	17.9
%RSD	.21199	.23455	.60311
#1	9165.3	55274.	2981.6
#2	9185.3	55069.	2950.7
#3	9204.2	55035.	2950.6

Approved: June 27, 2016

Sample Name: ICSAB Acquired: 6/25/2016 12:59:21 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51819**	**267.93**	**.25992**	**-.00749**	**.25012**	**.24499**	**247.43**	**.48653**
Stddev	.00274	.82	.00577	.00323	.00127	.00111	.96	.00069
%RSD	.52949	.30727	2.2185	43.201	.50926	.45301	.38893	.14255
#1	.51744	268.20	.26633	-.00904	.25065	.24547	247.63	.48716
#2	.52123	268.59	.25826	-.00965	.24866	.24578	246.38	.48579
#3	.51590	267.01	.25516	-.00377	.25104	.24372	248.28	.48665
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.24114**	**.25272**	**.25219**	**96.482**	**5.1931**	**.01181**	**249.25**	**.24440**
Stddev	.00029	.00219	.00111	.527	.0571	.00427	1.99	.00318
%RSD	.11831	.86807	.44033	.54597	1.0995	36.192	.80021	1.3017
#1	.24081	.25379	.25307	96.524	5.1297	.01134	249.32	.24432
#2	.24135	.25418	.25256	95.936	5.2405	.00779	247.22	.24763
#3	.24125	.25020	.25095	96.987	5.2091	.01630	251.20	.24127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00056**	**5.3241**	**.48364**	**.08508**	**.49426**	**.50665**	**.24857**	**.00554**
Stddev	.00038	.0216	.00155	.00195	.00579	.00649	.00790	.00251
%RSD	68.178	.40548	.32037	2.2888	1.1716	1.2807	3.1773	45.410
#1	-.00097	5.3151	.48290	.08463	.49996	.50736	.25719	.00434
#2	-.00022	5.3084	.48260	.08721	.48838	.49984	.24167	.00842
#3	-.00048	5.3487	.48542	.08340	.49443	.51276	.24687	.00384
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: ICSAB Acquired: 6/25/2016 12:59:21 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00018**	**.00065**	**-.00404**	**.46188**	**.25332**	**.48100**	**-.01374**
Stddev	.00129	.00010	.00503	.00318	.00107	.00194	.00218
%RSD	732.71	15.906	124.53	.68769	.42266	.40303	15.888
#1	.00146	.00072	-.00979	.45913	.25376	.48104	-.01122
#2	-.00112	.00053	-.00192	.46116	.25410	.48293	-.01498
#3	.00019	.00071	-.00042	.46536	.25210	.47905	-.01502
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9278.3**	**55833.**	**2985.5**
Stddev	15.8	186.	18.9
%RSD	.17001	.33320	.63177
#1	9263.8	55869.	2984.2
#2	9295.1	55632.	3004.9
#3	9275.9	55999.	2967.3

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 13:03:04 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.39792**	**10.019**	**.39868**	**.49351**	**.99429**	**.04982**	**9.9377**
Stddev	.00072	.017	.00144	.00412	.00753	.00017	.0996
%RSD	.18044	.17228	.36151	.83518	.75700	.33966	1.0022
#1	.39861	10.034	.39833	.49827	.98705	.04987	9.8876
#2	.39796	10.022	.40027	.49123	1.0021	.04997	10.052
#3	.39718	10.000	.39745	.49104	.99375	.04964	9.8731
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.04945**	**.20002**	**.50114**	**.50617**	**3.9950**	**49.191**	**.99914**
Stddev	.00013	.00034	.00115	.00138	.0385	.327	.00262
%RSD	.25449	.17079	.22943	.27257	.96478	.66521	.26204
#1	.04957	.20034	.49991	.50465	3.9625	48.889	.99615
#2	.04946	.20007	.50219	.50735	4.0376	49.538	1.0010
#3	.04932	.19966	.50132	.50650	3.9848	49.146	1.0003
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.9162**	**.49766**	**1.0071**	**49.316**	**.50460**	**9.9534**	**.50217**
Stddev	.1345	.00161	.0027	.281	.00141	.0056	.00090
%RSD	1.3564	.32258	.27182	.56936	.27993	.05661	.17861
#1	9.8004	.49634	1.0102	49.069	.50585	9.9507	.50282
#2	10.064	.49945	1.0061	49.622	.50307	9.9496	.50115
#3	9.8846	.49718	1.0051	49.257	.50488	9.9599	.50255
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 13:03:04 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1997**	**.40476**	**5.1642**	**1.0026**	**.99432**	**.99050**	**.50900**
Stddev	.0011	.00080	.0031	.0008	.00541	.00261	.00166
%RSD	.09538	.19644	.06047	.08057	.54374	.26324	.32573
#1	1.2008	.40408	5.1631	1.0035	.99081	.99348	.51044
#2	1.1985	.40563	5.1678	1.0020	1.0005	.98942	.50718
#3	1.1997	.40457	5.1618	1.0022	.99161	.98862	.50937
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0019**	**1.0053**	F **2.8691**
Stddev	.0036	.0009	.0111
%RSD	.35558	.09150	.38634
#1	.99784	1.0055	2.8628
#2	1.0045	1.0043	2.8819
#3	1.0033	1.0060	2.8625
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9877.8**	**59691.**	**3097.9**
Stddev	6.1	104.	11.8
%RSD	.06196	.17495	.38081
#1	9880.1	59766.	3105.5
#2	9870.8	59572.	3084.3
#3	9882.4	59736.	3103.8

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 13:06:41 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00082	.00013	.00057	.00151	.00063	.00002	-.01871	-.00022
Stddev	.00151	.00746	.00362	.00174	.00068	.00003	.02739	.00016
%RSD	185.23	5748.5	636.07	114.92	108.39	116.80	146.42	72.359
#1	.00201	.00304	-.00315	-.00049	.00141	.00001	-.04795	-.00036
#2	-.00088	-.00835	.00078	.00262	.00031	.00006	.00637	-.00027
#3	.00132	.00570	.00408	.00241	.00017	.00001	-.01455	-.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00002	.00054	.00076	.03608	.09214	-.00181	-.01859	-.00071
Stddev	.00025	.00025	.00080	.01975	.18930	.00302	.07756	.00052
%RSD	1039.3	46.422	105.73	54.731	205.46	166.60	417.27	72.248
#1	.00013	.00063	.00047	.01630	-.12492	-.00468	-.09449	-.00012
#2	.00011	.00026	.00167	.05580	.17827	.00135	-.02180	-.00101
#3	-.00032	.00073	.00014	.03615	.22306	-.00212	.06053	-.00101
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00274	.04401	-.00040	.00332	.00144	-.00005	.00482	-.00649
Stddev	.00039	.03341	.00024	.00918	.00111	.00350	.00394	.00014
%RSD	14.320	75.897	59.933	276.22	77.091	6590.6	81.757	2.1187
#1	.00273	.08068	-.00054	-.00157	.00018	-.00177	.00919	-.00633
#2	.00236	.03606	-.00012	.01392	.00189	-.00236	.00153	-.00657
#3	.00314	.01530	-.00055	-.00237	.00225	.00397	.00375	-.00656
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 13:06:41 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v974) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00049**	**.00039**	**.00186**	**-.00089**	**-.00002**	**.00003**	**.00447**
Stddev	.00029	.00029	.00050	.00195	.00109	.00004	.00129
%RSD	58.864	74.770	26.876	218.22	5660.3	163.33	28.933
#1	.00018	.00056	.00153	-.00246	.00118	.00003	.00395
#2	.00076	.00054	.00162	-.00151	-.00095	.00007	.00352
#3	.00054	.00005	.00244	.00129	-.00029	-.00002	.00594
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9714.8**	**59220.**	**3006.6**
Stddev	20.8	167.	18.8
%RSD	.21426	.28151	.62581
#1	9691.7	59319.	2985.4
#2	9732.2	59028.	3021.3
#3	9720.4	59314.	3013.1

Approved: June 27, 2016

Sample Name: PBW MC Acquired: 6/25/2016 13:18:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00038**	**.00511**	**.00086**	**-.00212**	**.00115**	**.00002**	**-.06747**	**.00005**
Stddev	.00209	.01093	.00117	.00067	.00097	.00011	.03960	.00035
%RSD	552.93	213.69	135.59	31.434	84.150	461.31	58.695	692.60
#1	-.00263	-.00536	-.00048	-.00211	.00011	.00001	-.09775	.00040
#2	.00149	.00426	.00140	-.00146	.00202	-.00008	-.02265	-.00029
#3	.00000	.01644	.00166	-.00279	.00133	.00014	-.08201	.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00010**	**-.00006**	**.00212**	**.02866**	**-.06431**	**.00509**	**.02259**	**-.00086**
Stddev	.00040	.00091	.00173	.01678	.07316	.00381	.05174	.00149
%RSD	400.23	1393.8	81.515	58.547	113.75	74.903	229.03	173.59
#1	.00052	-.00104	.00296	.02246	-.13473	.00394	-.01043	-.00102
#2	.00004	.00075	.00325	.01586	-.06951	.00199	-.00402	.00071
#3	-.00026	.00010	.00013	.04766	.01130	.00934	.08222	-.00227
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00087**	**-.00687**	**.00001**	**-.00017**	**.00005**	**-.00551**	**.00627**	**.00390**
Stddev	.00030	.03350	.00031	.00833	.00114	.00506	.00234	.00374
%RSD	34.257	487.68	3374.4	4899.3	2181.8	91.834	37.291	95.869
#1	.00091	.01218	.00037	.00506	.00058	-.00133	.00544	.00713
#2	.00056	-.04555	-.00014	-.00978	.00083	-.01113	.00446	-.00020
#3	.00115	.01276	-.00019	.00420	-.00126	-.00406	.00891	.00477
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: PBW MC Acquired: 6/25/2016 13:18:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00012**	**-.00022**	**-.00094**	**-.00120**	**.00059**	**.00019**	**.00136**
Stddev	.00090	.00014	.01001	.00401	.00016	.00009	.00212
%RSD	728.27	61.668	1066.1	333.15	27.157	48.058	155.52
#1	-.00022	-.00007	.00771	-.00405	.00041	.00016	-.00108
#2	.00115	-.00034	-.01190	.00338	.00070	.00012	.00241
#3	-.00056	-.00026	.00137	-.00294	.00066	.00029	.00275
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9802.9**	**59937.**	**3028.8**
Stddev	9.7	136.	16.1
%RSD	.09849	.22735	.53011
#1	9794.7	59785.	3046.3
#2	9813.5	59979.	3014.8
#3	9800.5	60047.	3025.3

Approved: June 27, 2016

Sample Name: LCSW MC Acquired: 6/25/2016 13:22:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19140**	**4.8040**	**.19266**	**.92897**	**.49049**	**.02396**	**4.9270**	**.02412**
Stddev	.00110	.0096	.00124	.00283	.00119	.00006	.0588	.00019
%RSD	.57611	.19942	.64179	.30493	.24273	.25817	1.1941	.78757
#1	.19253	4.7978	.19184	.92575	.49050	.02398	4.9121	.02430
#2	.19133	4.7992	.19408	.93105	.49168	.02402	4.8771	.02392
#3	.19033	4.8151	.19206	.93012	.48929	.02390	4.9919	.02415
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09725**	**.24390**	**.24732**	**1.9838**	**23.909**	**.49943**	**4.9211**	**.24209**
Stddev	.00025	.00056	.00120	.0346	.055	.00048	.0513	.00093
%RSD	.26181	.22973	.48504	1.7424	.23189	.09613	1.0429	.38463
#1	.09749	.24389	.24597	1.9470	23.904	.49915	4.9459	.24237
#2	.09699	.24447	.24774	1.9888	23.968	.49916	4.8621	.24284
#3	.09727	.24335	.24825	2.0156	23.857	.49999	4.9553	.24105
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49737**	**24.381**	**.24620**	**4.6982**	**.24508**	**.58185**	**.18662**	**2.4109**
Stddev	.00101	.058	.00194	.0026	.00506	.00206	.00438	.0033
%RSD	.20378	.23960	.78649	.05522	2.0646	.35379	2.3449	.13777
#1	.49803	24.349	.24767	4.6983	.25078	.58403	.18417	2.4135
#2	.49787	24.449	.24692	4.7007	.24112	.57994	.18402	2.4071
#3	.49620	24.346	.24401	4.6956	.24334	.58158	.19167	2.4120
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LCSW MC Acquired: 6/25/2016 13:22:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48537**	**.49494**	**.49102**	**.24360**	**.48594**	**.48236**	**.01763**
Stddev	.00248	.00157	.00916	.00175	.00116	.00127	.00304
%RSD	.51037	.31733	1.8663	.71814	.23927	.26228	17.214
#1	.48605	.49357	.49562	.24211	.48702	.48137	.01600
#2	.48744	.49666	.49698	.24315	.48610	.48194	.01576
#3	.48263	.49460	.48047	.24553	.48471	.48379	.02113
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9802.2**	**59979.**	**3063.3**
Stddev	16.1	75.	6.7
%RSD	.16429	.12520	.21884
#1	9786.8	59978.	3070.3
#2	9800.7	59904.	3056.9
#3	9818.9	60054.	3062.8

Approved: June 27, 2016

Sample Name: L1606080801 Acquired: 6/25/2016 13:25:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00120	-.00080	-.00163	.02599	.00717	.00000	50.564	.00016
Stddev	.00152	.00397	.00290	.00206	.00115	.00005	.112	.00037
%RSD	125.91	493.45	178.32	7.9153	16.060	1214.4	.22137	235.03
#1	.00226	.00203	-.00429	.02577	.00588	.00002	50.480	.00003
#2	-.00053	.00090	.00147	.02405	.00757	.00004	50.521	-.00014
#3	.00188	-.00534	-.00207	.02815	.00807	-.00005	50.691	.00058
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	.00083	.00107	.26409	1.2826	.00250	14.201	.11783
Stddev	.00018	.00044	.00061	.01139	.1290	.00132	.067	.00387
%RSD	127.01	53.307	57.172	4.3115	10.056	52.928	.47527	3.2872
#1	.00035	.00131	.00089	.27352	1.4315	.00167	14.162	.12070
#2	.00008	.00073	.00057	.25144	1.2130	.00180	14.162	.11343
#3	.00000	.00044	.00175	.26731	1.2035	.00402	14.279	.11937
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00063	15.281	.00045	.01127	-.00029	-.00250	-.00036	4.8502
Stddev	.00031	.048	.00109	.00496	.00257	.00086	.00870	.0263
%RSD	48.719	.31261	245.14	44.009	886.73	34.260	2404.6	.54311
#1	.00028	15.230	.00169	.00618	-.00269	-.00223	.00526	4.8717
#2	.00080	15.325	-.00034	.01609	.00243	-.00181	.00404	4.8581
#3	.00082	15.288	-.00001	.01154	-.00061	-.00346	-.01039	4.8208
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606080801 Acquired: 6/25/2016 13:25:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00081	.16531	-.00520	-.00297	.00071	.00304	.00248
Stddev	.00021	.00042	.00147	.00169	.00080	.00019	.00016
%RSD	26.039	.25638	28.334	56.894	114.04	6.1375	6.5372
#1	.00100	.16514	-.00627	-.00112	.00132	.00299	.00242
#2	.00058	.16499	-.00352	-.00336	-.00020	.00288	.00235
#3	.00084	.16579	-.00581	-.00443	.00100	.00324	.00266
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9767.4	60311.	3068.7
Stddev	25.3	182.	4.5
%RSD	.25865	.30239	.14809
#1	9787.8	60238.	3069.0
#2	9775.3	60177.	3064.1
#3	9739.2	60519.	3073.1

Approved: June 27, 2016

Sample Name: L1606080803S Acquired: 6/25/2016 13:29:50 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19425**	**4.8313**	**.19559**	**.97188**	**.49949**	**.02391**	**55.138**	**.02433**
Stddev	.00175	.0086	.00291	.00386	.00064	.00006	.286	.00001
%RSD	.89899	.17760	1.4870	.39759	.12859	.26078	.51899	.03982
#1	.19471	4.8260	.19892	.97122	.49908	.02384	54.823	.02434
#2	.19571	4.8412	.19357	.96839	.49916	.02394	55.383	.02434
#3	.19231	4.8266	.19428	.97603	.50023	.02395	55.206	.02432
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09623**	**.24025**	**.24291**	**2.2142**	**25.706**	**.49995**	**18.850**	**.36055**
Stddev	.00052	.00032	.00081	.0203	.056	.00310	.060	.00310
%RSD	.54129	.13263	.33315	.91450	.21668	.61973	.31584	.86069
#1	.09635	.23989	.24328	2.2335	25.718	.49651	18.866	.35746
#2	.09668	.24035	.24198	2.2161	25.755	.50082	18.784	.36367
#3	.09566	.24050	.24346	2.1931	25.645	.50252	18.899	.36052
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50273**	**39.946**	**.24148**	**4.8211**	**.24377**	**.58841**	**.19283**	**7.2475**
Stddev	.00069	.089	.00226	.0078	.00185	.00329	.00168	.0253
%RSD	.13674	.22372	.93733	.16218	.76094	.55837	.87085	.34834
#1	.50346	39.846	.24394	4.8298	.24591	.59218	.19095	7.2767
#2	.50266	40.018	.24099	4.8187	.24276	.58614	.19336	7.2326
#3	.50209	39.974	.23950	4.8148	.24264	.58690	.19419	7.2333
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606080803S Acquired: 6/25/2016 13:29:50 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-07

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48600**	**.66572**	**.49120**	**.24058**	**.48625**	**.48121**	**.00549**
Stddev	.00062	.00113	.00488	.00183	.00111	.00310	.00062
%RSD	.12721	.16988	.99375	.75886	.22783	.64511	11.331
#1	.48664	.66443	.49684	.23968	.48515	.48355	.00538
#2	.48596	.66654	.48839	.24268	.48622	.48240	.00493
#3	.48540	.66619	.48838	.23938	.48737	.47769	.00616
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9640.1**	**59155.**	**3051.9**
Stddev	28.9	91.	4.4
%RSD	.29949	.15374	.14285
#1	9607.9	59063.	3056.9
#2	9648.6	59158.	3049.0
#3	9663.8	59244.	3049.8

Approved: June 27, 2016

Sample Name: L1606080804SD Acquired: 6/25/2016 13:33:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.19367	4.8696	.19395	.97468	.50079	.02428	55.592	.02458
Stddev	.00103	.0189	.00351	.00126	.00119	.00004	.087	.00002
%RSD	.53221	.38806	1.8082	.12977	.23838	.15828	.15705	.06402
#1	.19313	4.8871	.19259	.97422	.49995	.02427	55.687	.02459
#2	.19486	4.8723	.19133	.97371	.50216	.02425	55.574	.02456
#3	.19302	4.8496	.19794	.97611	.50027	.02433	55.515	.02458
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09687	.24382	.24330	2.1942	25.602	.50391	18.927	.35670
Stddev	.00034	.00139	.00044	.0216	.187	.00317	.145	.00440
%RSD	.35450	.56820	.18086	.98630	.73073	.62836	.76565	1.2346
#1	.09649	.24542	.24364	2.2182	25.415	.50090	18.778	.36166
#2	.09698	.24295	.24347	2.1881	25.600	.50721	19.067	.35519
#3	.09715	.24309	.24280	2.1762	25.789	.50361	18.935	.35325
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.50687	40.011	.24239	4.8324	.24609	.59184	.19199	7.3269
Stddev	.00174	.019	.00180	.0039	.00113	.00410	.00494	.0108
%RSD	.34421	.04769	.74422	.07995	.45954	.69206	2.5730	.14791
#1	.50863	40.009	.24329	4.8354	.24728	.59240	.18629	7.3276
#2	.50684	40.031	.24358	4.8338	.24503	.59563	.19457	7.3373
#3	.50514	39.993	.24032	4.8281	.24594	.58749	.19509	7.3157
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606080804SD Acquired: 6/25/2016 13:33:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573190-08

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.48884	.66564	.48806	.24641	.49385	.48167	.00481
Stddev	.00243	.00060	.00210	.00100	.00262	.00319	.00105
%RSD	.49731	.09007	.43005	.40687	.53028	.66320	21.781
#1	.48902	.66631	.48604	.24752	.49576	.48059	.00363
#2	.49118	.66515	.49023	.24612	.49493	.48526	.00518
#3	.48633	.66545	.48792	.24558	.49086	.47916	.00563
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9643.8	58961.	3032.4
Stddev	19.6	179.	14.4
%RSD	.20367	.30279	.47352
#1	9664.0	58941.	3022.3
#2	9624.8	58793.	3026.2
#3	9642.5	59149.	3048.8

Approved: June 27, 2016

Sample Name: L1606082203 Acquired: 6/25/2016 13:37:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00071	.03250	.00108	.03413	.76881	.00001	31.323	.00009
Stddev	.00326	.00113	.00298	.00231	.00291	.00007	.069	.00003
%RSD	461.08	3.4738	274.98	6.7818	.37861	986.76	.21967	35.456
#1	-.00056	.03174	-.00033	.03677	.76753	-.00000	31.323	.00006
#2	.00440	.03379	-.00093	.03317	.76676	.00008	31.255	.00011
#3	-.00173	.03195	.00451	.03246	.77215	-.00006	31.393	.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00298	.00799	.00175	5.1531	49.440	.10315	14.307	.17132
Stddev	.00036	.00030	.00156	.0220	.158	.00201	.082	.00563
%RSD	12.200	3.7657	89.202	.42624	.31871	1.9519	.57045	3.2873
#1	.00316	.00823	.00355	5.1330	49.302	.10173	14.214	.17528
#2	.00323	.00810	.00078	5.1498	49.405	.10227	14.338	.17381
#3	.00256	.00765	.00092	5.1766	49.612	.10546	14.368	.16487
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00045	90.168	.00369	.00788	-.00122	.00230	.00147	11.287
Stddev	.00023	.210	.00039	.00442	.00289	.00119	.00467	.027
%RSD	50.114	.23246	10.480	56.105	237.50	51.496	317.01	.23968
#1	.00031	90.009	.00403	.01267	.00108	.00337	.00397	11.318
#2	.00033	90.090	.00327	.00394	-.00447	.00102	-.00391	11.275
#3	.00071	90.405	.00377	.00704	-.00027	.00252	.00436	11.268
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082203 Acquired: 6/25/2016 13:37:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00007**	**.79119**	**.00045**	**-.00075**	**.00043**	**.00534**	**-.00006**
Stddev	.00030	.00071	.00315	.00229	.00033	.00006	.00108
%RSD	437.99	.09029	691.72	303.10	75.619	1.1956	1667.5
#1	-.00035	.79036	-.00147	.00135	.00018	.00529	-.00108
#2	-.00009	.79165	-.00125	-.00319	.00032	.00542	.00108
#3	.00024	.79155	.00408	-.00042	.00080	.00532	-.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9649.6**	**58766.**	**3005.5**
Stddev	29.5	106.	1.2
%RSD	.30621	.18076	.04053
#1	9617.1	58644.	3006.7
#2	9657.0	58831.	3005.5
#3	9674.8	58825.	3004.3

Approved: June 27, 2016

Sample Name: L1606082202 Acquired: 6/25/2016 13:40:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573973-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00163**	**.00700**	**-.00255**	**.09063**	**.14584**	**.00010**	**6.2202**	**-.00011**
Stddev	.00068	.00493	.00215	.00172	.00146	.00002	.0817	.00004
%RSD	41.927	70.458	84.324	1.8967	.99792	21.043	1.3126	39.110
#1	.00225	.01016	-.00441	.09032	.14618	.00008	6.1515	-.00007
#2	.00090	.00132	-.00302	.08908	.14425	.00013	6.1986	-.00016
#3	.00175	.00952	-.00020	.09248	.14710	.00010	6.3104	-.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00004**	**.00096**	**.00077**	**2.0756**	**1.5049**	**.02492**	**4.3992**	**.07622**
Stddev	.00041	.00002	.00085	.0056	.1629	.00329	.1257	.00201
%RSD	1133.0	2.0322	110.52	.26897	10.822	13.197	2.8567	2.6344
#1	-.00024	.00097	.00165	2.0692	1.6925	.02131	4.4123	.07724
#2	.00043	.00097	.00070	2.0781	1.3995	.02572	4.2675	.07752
#3	-.00030	.00094	-.00004	2.0795	1.4227	.02774	4.5178	.07391
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00024**	**83.246**	**.00061**	**.10079**	**.00421**	**-.00520**	**.00452**	**12.135**
Stddev	.00009	.263	.00086	.00636	.00131	.00271	.00174	.029
%RSD	36.387	.31568	140.09	6.3065	31.208	52.125	38.464	.24257
#1	-.00033	83.275	.00060	.10060	.00298	-.00301	.00254	12.154
#2	-.00015	82.970	.00147	.10724	.00559	-.00823	.00579	12.151
#3	-.00025	83.493	-.00024	.09454	.00404	-.00437	.00523	12.101
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082202 Acquired: 6/25/2016 13:40:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573973-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00029	.21054	.00087	-.00133	-.00012	.00182	.00052
Stddev	.00041	.00091	.00263	.00076	.00048	.00007	.00152
%RSD	145.47	.43126	300.38	57.105	397.63	3.7016	294.24
#1	.00007	.21002	.00237	-.00045	.00028	.00184	-.00094
#2	.00003	.21002	.00241	-.00174	.00001	.00175	.00210
#3	.00076	.21159	-.00216	-.00179	-.00065	.00188	.00040
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9729.1	59376.	3030.4
Stddev	8.6	260.	19.9
%RSD	.08835	.43738	.65516
#1	9720.3	59079.	3018.5
#2	9729.3	59490.	3053.3
#3	9737.5	59559.	3019.5

Approved: June 27, 2016

Sample Name: L1606082202PS Acquired: 6/25/2016 13:44:52 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19375**	**4.8752**	**.19613**	**1.0291**	**.62833**	**.02428**	**10.648**	**.02477**
Stddev	.00051	.0099	.00160	.0019	.00588	.00006	.170	.00011
%RSD	.26273	.20232	.81334	.18131	.93614	.24036	1.5959	.45206
#1	.19417	4.8642	.19702	1.0298	.63084	.02431	10.660	.02468
#2	.19391	4.8782	.19429	1.0305	.63254	.02431	10.812	.02489
#3	.19319	4.8832	.19709	1.0270	.62161	.02421	10.473	.02473
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09829**	**.24324**	**.24855**	**3.8486**	**25.821**	**.52335**	**8.7582**	**.31225**
Stddev	.00082	.00053	.00056	.0314	.285	.00671	.0434	.00355
%RSD	.83225	.21842	.22566	.81716	1.1038	1.2812	.49530	1.1380
#1	.09742	.24374	.24849	3.8714	25.861	.52361	8.7634	.31338
#2	.09905	.24330	.24914	3.8618	26.083	.52993	8.7125	.31510
#3	.09841	.24268	.24802	3.8128	25.517	.51653	8.7988	.30827
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50486**	**99.997**	**.24562**	**4.8721**	**.24648**	**.59017**	**.19243**	**13.471**
Stddev	.00060	.660	.00153	.0054	.00424	.00393	.00054	.013
%RSD	.11858	.66045	.62190	.11137	1.7208	.66559	.28037	.09301
#1	.50422	100.35	.24553	4.8783	.24168	.58648	.19181	13.458
#2	.50541	100.41	.24414	4.8682	.24971	.59430	.19267	13.471
#3	.50494	99.235	.24719	4.8698	.24807	.58975	.19281	13.483
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082202PS Acquired: 6/25/2016 13:44:52 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48807**	**.68216**	**.50019**	**.24323**	**.49372**	**.48957**	**.00258**
Stddev	.00205	.00533	.01089	.00117	.00097	.00018	.00106
%RSD	.41999	.78112	2.1777	.48173	.19621	.03662	41.003
#1	.49029	.68537	.51266	.24188	.49454	.48951	.00281
#2	.48770	.68509	.49539	.24396	.49265	.48978	.00349
#3	.48624	.67601	.49252	.24385	.49397	.48943	.00142
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9652.9**	**58738.**	**3022.5**
Stddev	16.4	44.	32.4
%RSD	.17031	.07498	1.0716
#1	9668.7	58757.	3005.1
#2	9654.1	58688.	3002.5
#3	9635.9	58769.	3059.8

Approved: June 27, 2016

Sample Name: L1606082202SDL Acquired: 6/25/2016 13:48:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00041	-.00874	.00232	.01966	.02846	.00012	1.1981	-.00020
Stddev	.00102	.00520	.00129	.00166	.00164	.00003	.0364	.00029
%RSD	249.96	59.552	55.713	8.4558	5.7787	28.080	3.0406	145.60
#1	-.00049	-.00387	.00244	.02114	.03034	.00013	1.2234	-.00054
#2	.00065	-.00812	.00098	.01996	.02778	.00008	1.1564	-.00008
#3	-.00139	-.01423	.00356	.01786	.02727	.00014	1.2146	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00018	.00097	.00049	.42314	.31666	.00414	.84994	.01478
Stddev	.00036	.00092	.00031	.02269	.03395	.00078	.10125	.00159
%RSD	202.36	94.832	64.039	5.3615	10.720	18.849	11.912	10.770
#1	-.00059	-.00007	.00031	.40193	.27970	.00483	.89447	.01388
#2	-.00002	.00128	.00031	.42043	.32384	.00429	.92129	.01385
#3	.00008	.00169	.00086	.44706	.34645	.00330	.73406	.01662
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00015	16.797	-.00010	.02573	.00379	-.00654	-.00062	2.4606
Stddev	.00016	.116	.00075	.00347	.00211	.00206	.00121	.0014
%RSD	105.18	.68795	766.63	13.470	55.760	31.494	195.90	.05654
#1	-.00025	16.870	-.00048	.02714	.00159	-.00786	-.00167	2.4590
#2	-.00024	16.856	-.00058	.02827	.00580	-.00417	-.00090	2.4616
#3	.00003	16.663	.00077	.02178	.00399	-.00761	.00071	2.4611
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082202SDL Acquired: 6/25/2016 13:48:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00034**	**.04233**	**.00225**	**-.00298**	**.00005**	**.00069**	**.00014**
Stddev	.00035	.00076	.00380	.00312	.00068	.00008	.00023
%RSD	102.54	1.7878	169.10	104.78	1298.3	11.967	160.88
#1	-.00033	.04309	.00306	.00056	-.00034	.00074	-.00011
#2	.00000	.04232	.00558	-.00534	-.00035	.00059	.00032
#3	-.00071	.04157	-.00189	-.00415	.00084	.00073	.00022
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10112.**	**61959.**	**3102.5**
Stddev	25.	335.	21.2
%RSD	.24804	.54051	.68302
#1	10140.	62322.	3081.9
#2	10091.	61890.	3101.4
#3	10107.	61663.	3124.2

Approved: June 27, 2016

Sample Name: L1606082202SDL Acquired: 6/25/2016 13:52:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00014**	**-.00201**	**.00162**	**.00506**	**.00674**	**.00009**	**.22194**	**.00006**
Stddev	.00098	.00612	.00099	.00223	.00036	.00007	.03121	.00010
%RSD	701.06	303.84	61.332	44.136	5.3564	71.933	14.062	159.17
#1	-.00091	-.00825	.00149	.00575	.00715	.00008	.25227	.00003
#2	-.00048	-.00177	.00268	.00256	.00663	.00004	.22363	-.00001
#3	.00097	.00398	.00070	.00686	.00645	.00017	.18992	.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00007**	**.00003**	**.00093**	**.10294**	**.07211**	**-.00053**	**.11677**	**.00303**
Stddev	.00044	.00176	.00026	.00844	.06933	.00190	.04489	.00266
%RSD	660.12	5191.6	27.598	8.2017	96.144	360.65	38.441	87.562
#1	-.00052	.00076	.00092	.11264	.12436	-.00025	.12873	.00344
#2	.00037	-.00197	.00068	.09726	-.00654	-.00255	.06712	.00020
#3	-.00005	.00131	.00120	.09891	.09851	.00122	.15447	.00546
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00046**	**3.4626**	**-.00095**	**.01102**	**.00309**	**-.00448**	**.00254**	**.52285**
Stddev	.00009	.0579	.00053	.00740	.00233	.00663	.00646	.00911
%RSD	18.881	1.6714	55.686	67.190	75.440	147.88	254.92	1.7427
#1	-.00041	3.4180	-.00056	.01783	.00042	-.00172	-.00323	.51493
#2	-.00042	3.4418	-.00074	.01207	.00473	-.01205	.00952	.52082
#3	-.00056	3.5280	-.00155	.00314	.00412	.00032	.00131	.53281
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082202SDL Acquired: 6/25/2016 13:52:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573973-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00039**	**.00876**	**-.00160**	**-.00170**	**.00064**	**.00048**	**.00025**
Stddev	.00024	.00020	.00243	.00072	.00151	.00001	.00052
%RSD	61.212	2.3382	151.62	42.632	234.58	1.6034	205.72
#1	-.00036	.00876	-.00015	-.00219	-.00097	.00048	.00032
#2	-.00064	.00897	-.00025	-.00087	.00088	.00049	-.00030
#3	-.00017	.00856	-.00441	-.00204	.00202	.00048	.00073
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10010.**	**61091.**	**3050.7**
Stddev	14.	94.	16.7
%RSD	.14368	.15455	.54708
#1	9999.9	61099.	3067.0
#2	10027.	60993.	3033.7
#3	10004.	61182.	3051.5

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 13:56:22 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39703	9.8823	.39208	.48360	.99649	.04877	9.9442
Stddev	.00097	.0185	.00108	.00155	.00485	.00012	.0391
%RSD	.24464	.18766	.27635	.32015	.48698	.23806	.39293
#1	.39651	9.9033	.39094	.48432	.99107	.04885	9.8991
#2	.39815	9.8682	.39222	.48182	1.0004	.04883	9.9670
#3	.39642	9.8754	.39309	.48466	.99796	.04864	9.9665
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04952	.19590	.48861	.49872	3.9553	49.086	.99832
Stddev	.00026	.00028	.00039	.00131	.0110	.181	.00236
%RSD	.52554	.14170	.07934	.26171	.27922	.36920	.23624
#1	.04970	.19602	.48888	.50017	3.9554	48.878	.99571
#2	.04922	.19610	.48816	.49764	3.9662	49.210	1.0003
#3	.04964	.19559	.48877	.49836	3.9441	49.170	.99895
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.6989	.49966	.99513	49.378	.49577	9.7713	.49937
Stddev	.0286	.00128	.00487	.189	.00112	.0077	.00125
%RSD	.29470	.25687	.48908	.38358	.22653	.07875	.25024
#1	9.6936	.49933	.99926	49.226	.49626	9.7672	.49899
#2	9.6733	.50107	.99636	49.590	.49656	9.7802	.50076
#3	9.7297	.49857	.98976	49.317	.49448	9.7665	.49835
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 13:56:22 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.1744	.39516	5.0749	.97934	.99997	.99747	.49995
Stddev	.0050	.00337	.0045	.00061	.00261	.00842	.00315
%RSD	.42778	.85272	.08873	.06267	.26098	.84451	.63052
#1	1.1693	.39868	5.0732	.98005	.99777	.99280	.50104
#2	1.1793	.39197	5.0800	.97900	1.0028	1.0072	.49639
#3	1.1748	.39484	5.0715	.97897	.99928	.99242	.50240
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.98727	.98962	F 2.8790
Stddev	.00224	.00122	.0125
%RSD	.22699	.12332	.43408
#1	.98696	.98924	2.8649
#2	.98519	.99098	2.8888
#3	.98965	.98864	2.8834
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9881.8	60025.	3051.6
Stddev	7.5	117.	13.0
%RSD	.07613	.19437	.42745
#1	9874.3	60038.	3066.6
#2	9889.3	60134.	3043.0
#3	9881.9	59902.	3045.3

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 14:00:01 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00037	.00384	.00171	.00288	.00164	.00016	-.02160	-.00002
Stddev	.00056	.00607	.00179	.00157	.00101	.00008	.03268	.00008
%RSD	153.00	158.06	104.81	54.496	61.686	49.488	151.32	357.31
#1	-.00082	.00133	.00363	.00341	.00099	.00008	-.05876	.00002
#2	-.00054	.01077	.00142	.00412	.00280	.00016	-.00868	-.00011
#3	.00026	-.00057	.00008	.00112	.00111	.00023	.00265	.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	.00089	.00145	.02539	.05830	-.00178	-.00516	-.00289
Stddev	.00025	.00059	.00085	.02878	.16368	.00331	.04118	.00267
%RSD	184.36	65.782	58.421	113.33	280.73	185.62	797.74	92.366
#1	-.00007	.00150	.00078	.04856	-.06991	-.00470	-.03370	-.00034
#2	.00042	.00086	.00240	.03444	.24267	-.00246	.04205	-.00265
#3	.00006	.00033	.00117	-.00682	.00216	.00181	-.02383	-.00566
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00312	.01750	-.00036	.00463	.00110	.00408	.00171	-.00349
Stddev	.00035	.01788	.00074	.00054	.00119	.00430	.00409	.00216
%RSD	11.208	102.15	207.33	11.581	107.78	105.36	238.51	61.802
#1	.00276	.00856	.00044	.00497	-.00027	-.00057	.00622	-.00129
#2	.00346	.03809	-.00047	.00401	.00175	.00791	-.00177	-.00560
#3	.00314	.00586	-.00104	.00491	.00181	.00489	.00070	-.00357
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 14:00:01 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00015	.00055	-.00034	-.00341	.00088	.00013	.00321
Stddev	.00088	.00013	.00661	.00148	.00029	.00021	.00123
%RSD	577.33	24.369	1964.7	43.438	32.521	159.30	38.279
#1	-.00076	.00044	.00453	-.00437	.00106	-.00009	.00194
#2	.00100	.00070	.00232	-.00416	.00055	.00032	.00329
#3	.00022	.00052	-.00786	-.00171	.00104	.00016	.00439
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9674.1	59179.	2989.6
Stddev	10.8	80.	18.0
%RSD	.11204	.13474	.60268
#1	9680.4	59116.	2972.7
#2	9661.6	59269.	3008.6
#3	9680.2	59153.	2987.4

Approved: June 27, 2016

Sample Name: L1606082302 Acquired: 6/25/2016 14:03:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00010**	**.01235**	**.00398**	**.02065**	**7.6114**	**.00014**	**242.99**
Stddev	.00016	.00738	.00092	.00248	.0086	.00009	.50
%RSD	156.01	59.723	23.024	12.009	.11234	64.318	.20416
#1	-.00023	.01740	.00411	.01789	7.6202	.00016	242.85
#2	-.00014	.01576	.00483	.02137	7.6109	.00023	243.54
#3	.00007	.00389	.00301	.02270	7.6032	.00004	242.58
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00045**	**.05380**	**.00122**	**.00261**	**4.4999**	**3.4068**	**.13022**
Stddev	.00043	.00087	.00035	.00011	.0119	.0234	.00431
%RSD	95.240	1.6084	28.402	4.2649	.26410	.68699	3.3105
#1	-.00003	.05465	.00142	.00253	4.4869	3.3837	.13493
#2	.00060	.05384	.00143	.00274	4.5026	3.4062	.12925
#3	.00078	.05292	.00082	.00256	4.5101	3.4305	.12648
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**186.55**	**.80417**	**.00043**	F **676.67**	**.01798**	**.04840**	**.00356**
Stddev	.43	.00344	.00029	5.78	.00081	.00767	.00232
%RSD	.22901	.42830	67.402	.85481	4.4962	15.839	65.024
#1	186.94	.80321	.00050	676.18	.01876	.05723	.00107
#2	186.09	.80800	.00067	682.68	.01714	.04349	.00566
#3	186.63	.80131	.00011	671.15	.01803	.04447	.00396
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 27, 2016

Sample Name: L1606082302 Acquired: 6/25/2016 14:03:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00329**	**-.00499**	**17.541**	**.00002**	**5.8024**	F **-.03279**	**-.00373**
Stddev	.00417	.00266	.115	.00016	.0077	.00249	.00217
%RSD	126.77	53.331	.65409	771.83	.13255	7.6037	58.233
#1	-.00765	-.00207	17.622	-.00016	5.8086	-.03219	-.00297
#2	-.00289	-.00563	17.591	.00008	5.8048	-.03065	-.00619
#3	.00066	-.00727	17.409	.00014	5.7938	-.03553	-.00205
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00105**	**.00822**	**.00588**
Stddev	.00081	.00012	.00100
%RSD	77.437	1.5051	16.931
#1	-.00020	.00826	.00695
#2	-.00112	.00832	.00498
#3	-.00182	.00808	.00571
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9001.2**	**54623.**	**2943.0**
Stddev	12.7	293.	11.3
%RSD	.14121	.53595	.38250
#1	8986.8	54359.	2937.9
#2	9010.8	54574.	2935.3
#3	9006.1	54938.	2955.9

Approved: June 27, 2016

Sample Name: L1606082303 Acquired: 6/25/2016 14:07:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00097**	**.07689**	**.00082**	**.01317**	**4.0611**	**.00008**	**194.35**
Stddev	.00087	.00794	.00214	.00140	.0013	.00002	.19
%RSD	89.857	10.324	259.83	10.655	.03277	29.564	.09947
#1	.00177	.08422	.00140	.01160	4.0625	.00007	194.28
#2	.00005	.07798	.00261	.01363	4.0599	.00006	194.19
#3	.00109	.06846	-.00155	.01429	4.0610	.00010	194.56
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00041**	**.00720**	**.00128**	**.00172**	**.48254**	**1.5553**	**.05921**
Stddev	.00016	.00037	.00008	.00037	.03011	.0518	.00214
%RSD	37.982	5.1555	6.4501	21.255	6.2390	3.3303	3.6194
#1	.00025	.00709	.00125	.00159	.48289	1.5549	.05878
#2	.00056	.00690	.00137	.00144	.45226	1.5037	.05732
#3	.00044	.00762	.00121	.00214	.51247	1.6073	.06154
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**110.52**	**2.3053**	**-.00008**	F **614.40**	**.00312**	**.00621**	**.00333**
Stddev	.38	.0070	.00038	4.80	.00050	.00738	.00101
%RSD	.34277	.30352	457.85	.78181	15.943	118.75	30.264
#1	110.96	2.3090	-.00052	609.47	.00307	-.00184	.00218
#2	110.30	2.2972	.00010	614.65	.00363	.01264	.00406
#3	110.30	2.3096	.00018	619.07	.00264	.00783	.00375
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 27, 2016

Sample Name: L1606082303 Acquired: 6/25/2016 14:07:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00131	-.00398	12.572	-.00061	3.7282	-.02186	-.00195
Stddev	.00101	.00138	.069	.00058	.0049	.00737	.00114
%RSD	77.475	34.690	.54565	95.484	.12996	33.713	58.500
#1	-.00178	-.00242	12.599	.00005	3.7247	-.01374	-.00282
#2	-.00014	-.00503	12.623	-.00084	3.7262	-.02372	-.00066
#3	-.00200	-.00451	12.494	-.00105	3.7337	-.02812	-.00237
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	-.00017	.00519	.00587
Stddev	.00034	.00017	.00146
%RSD	196.61	3.3392	24.813
#1	.00020	.00538	.00454
#2	-.00026	.00515	.00565
#3	-.00046	.00504	.00743
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9113.2	55170.	2962.5
Stddev	17.2	284.	10.8
%RSD	.18895	.51491	.36572
#1	9120.9	54953.	2952.0
#2	9093.5	55065.	2973.6
#3	9125.3	55492.	2961.9

Approved: June 27, 2016

Sample Name: L1606082304 Acquired: 6/25/2016 14:12:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00108**	**.00282**	**.00375**	**.01195**	**1.2873**	**.00004**	**44.812**	**-.00032**
Stddev	.00081	.00175	.00233	.00191	.0068	.00003	.140	.00033
%RSD	75.442	62.066	62.147	15.958	.52861	68.482	.31190	102.13
#1	-.00032	.00374	.00333	.01352	1.2889	.00004	44.884	-.00060
#2	-.00194	.00392	.00166	.00983	1.2799	.00001	44.651	.00004
#3	-.00098	.00080	.00626	.01250	1.2932	.00007	44.900	-.00039
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00062**	**-.00235**	**.00403**	**80.696**	**3.0351**	**.04425**	**27.190**	**1.4212**
Stddev	.00024	.00090	.00186	.288	.0674	.00194	.251	.0124
%RSD	38.213	38.227	46.192	.35704	2.2197	4.3802	.92249	.86964
#1	.00059	-.00132	.00534	80.883	2.9579	.04406	26.967	1.4224
#2	.00088	-.00276	.00485	80.364	3.0650	.04628	27.141	1.4083
#3	.00040	-.00297	.00190	80.840	3.0823	.04242	27.462	1.4329
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00083**	**136.40**	**.00117**	**.05473**	**.00702**	**-.00148**	**-.00406**	**16.093**
Stddev	.00029	.62	.00056	.00919	.00288	.00326	.00473	.052
%RSD	35.125	.45656	47.701	16.792	41.072	220.80	116.61	.32014
#1	-.00064	136.42	.00054	.06475	.00887	-.00465	-.00925	16.145
#2	-.00116	135.77	.00135	.05277	.00848	.00187	.00003	16.091
#3	-.00068	137.01	.00161	.04668	.00370	-.00165	-.00296	16.042
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606082304 Acquired: 6/25/2016 14:12:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00141**	**1.0265**	**.02088**	**-.00129**	**-.00025**	**.01157**	**-.01042**
Stddev	.00052	.0054	.00601	.00283	.00173	.00020	.00036
%RSD	36.987	.52659	28.762	219.55	704.92	1.7611	3.4676
#1	.00201	1.0281	.02616	-.00374	-.00225	.01176	-.01080
#2	.00117	1.0205	.01435	.00181	.00074	.01159	-.01039
#3	.00105	1.0309	.02213	-.00193	.00077	.01136	-.01008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9584.5**	**58328.**	**2992.4**
Stddev	11.8	87.	5.0
%RSD	.12317	.14881	.16820
#1	9573.1	58228.	2986.6
#2	9583.6	58387.	2995.2
#3	9596.7	58368.	2995.4

Approved: June 27, 2016

Sample Name: L1606114801 Acquired: 6/25/2016 14:15:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00263	.73135	.00385	.12210	.08262	.00020	27.646	.00018
Stddev	.00073	.00428	.00128	.00118	.00156	.00004	.096	.00015
%RSD	27.764	.58467	33.324	.96789	1.8866	19.880	.34678	80.450
#1	.00220	.73627	.00263	.12346	.08403	.00016	27.689	.00009
#2	.00223	.72927	.00372	.12147	.08287	.00024	27.713	.00035
#3	.00348	.72851	.00519	.12136	.08095	.00020	27.536	.00011
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00054	.00138	.00319	.50952	4.5430	.00820	13.834	.02681
Stddev	.00058	.00089	.00080	.00314	.0592	.00406	.158	.00190
%RSD	107.78	64.361	25.113	.61565	1.3042	49.473	1.1398	7.0691
#1	-.00010	.00066	.00412	.50590	4.4785	.00685	13.879	.02807
#2	.00067	.00110	.00265	.51133	4.5951	.01276	13.964	.02463
#3	.00103	.00237	.00282	.51133	4.5554	.00499	13.658	.02774
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00124	65.018	.00157	.05465	.00287	-.00148	.00632	4.3760
Stddev	.00020	.232	.00022	.00221	.00233	.00288	.00218	.0207
%RSD	16.215	.35629	13.957	4.0445	81.267	194.34	34.507	.47244
#1	.00141	65.009	.00133	.05367	.00373	-.00288	.00878	4.3946
#2	.00130	65.254	.00176	.05719	.00465	-.00339	.00555	4.3795
#3	.00102	64.791	.00160	.05310	.00023	.00183	.00463	4.3537
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606114801 Acquired: 6/25/2016 14:15:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00055	.26681	.01202	-.00269	.00480	.00516	.00567
Stddev	.00125	.00125	.00081	.00070	.00038	.00011	.00071
%RSD	227.21	.46986	6.7372	25.927	7.9446	2.0641	12.451
#1	-.00072	.26656	.01275	-.00259	.00452	.00521	.00488
#2	.00178	.26818	.01115	-.00205	.00524	.00504	.00622
#3	.00058	.26571	.01216	-.00344	.00465	.00524	.00592
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9685.5	59237.	3013.9
Stddev	15.3	70.	21.8
%RSD	.15804	.11840	.72168
#1	9667.9	59212.	3004.6
#2	9692.8	59182.	2998.3
#3	9695.8	59316.	3038.7

Approved: June 27, 2016

Sample Name: L1606114802 Acquired: 6/25/2016 14:19:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00086	.02576	.00392	.13054	.14235	.00004	53.032	.00013
Stddev	.00161	.00264	.00128	.00304	.00093	.00006	.088	.00011
%RSD	186.85	10.249	32.777	2.3320	.65273	165.85	.16643	81.772
#1	-.00048	.02479	.00505	.13055	.14272	-.00002	52.942	.00017
#2	.00053	.02374	.00418	.13357	.14130	.00011	53.037	.00022
#3	-.00263	.02875	.00252	.12748	.14305	.00003	53.119	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00081	.00416	.00255	.05069	2.7172	.00979	45.232	.00619
Stddev	.00057	.00090	.00135	.00895	.0969	.00331	.213	.00128
%RSD	70.201	21.666	53.148	17.648	3.5679	33.803	.47142	20.757
#1	.00109	.00488	.00157	.05068	2.6091	.01132	44.992	.00715
#2	.00016	.00315	.00198	.05964	2.7964	.01206	45.307	.00668
#3	.00119	.00444	.00409	.04175	2.7463	.00599	45.398	.00473
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	113.30	.00198	.03363	.00397	-.00032	.00108	5.7426
Stddev	.00005	.15	.00097	.00220	.00496	.00242	.00246	.0270
%RSD	9.2544	.13423	48.767	6.5321	124.89	765.53	227.39	.46957
#1	.00062	113.13	.00098	.03204	.00115	-.00095	-.00128	5.7131
#2	.00062	113.39	.00206	.03614	.00107	.00236	.00363	5.7660
#3	.00052	113.39	.00290	.03272	.00971	-.00236	.00090	5.7486
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606114802 Acquired: 6/25/2016 14:19:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00015**	**.36906**	**-.00370**	**-.00072**	**.00779**	**.00692**	**.00331**
Stddev	.00028	.00064	.00411	.00265	.00075	.00021	.00138
%RSD	191.03	.17400	110.98	366.07	9.6620	3.0569	41.716
#1	-.00017	.36882	.00101	-.00049	.00853	.00671	.00489
#2	.00034	.36857	-.00557	.00180	.00703	.00713	.00270
#3	.00027	.36978	-.00654	-.00348	.00781	.00694	.00235
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9627.8**	**58951.**	**3062.4**
Stddev	26.1	26.	1.9
%RSD	.27151	.04447	.06237
#1	9657.6	58927.	3060.7
#2	9617.0	58947.	3062.0
#3	9608.7	58979.	3064.5

Approved: June 27, 2016

Sample Name: L1606115001 Acquired: 6/25/2016 14:23:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00134	.94860	.00623	.13672	.10593	.00014	35.066	.00002
Stddev	.00163	.01243	.00298	.00233	.00120	.00009	.208	.00013
%RSD	122.05	1.3105	47.819	1.7071	1.1370	65.235	.59262	741.62
#1	-.00020	.96246	.00332	.13853	.10456	.00018	34.826	.00011
#2	.00116	.93843	.00928	.13754	.10682	.00004	35.168	-.00013
#3	.00306	.94490	.00609	.13408	.10640	.00021	35.202	.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00095	.00115	.00402	.75423	2.6231	.00517	21.361	.05361
Stddev	.00038	.00077	.00076	.00983	.0329	.00134	.091	.00159
%RSD	39.586	67.102	19.015	1.3028	1.2541	25.927	.42723	2.9705
#1	.00055	.00114	.00333	.74497	2.6490	.00636	21.310	.05382
#2	.00102	.00192	.00389	.75319	2.6342	.00372	21.307	.05192
#3	.00129	.00038	.00484	.76454	2.5861	.00543	21.467	.05508
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00320	114.16	.00202	.11444	.00566	-.00032	.00352	5.0408
Stddev	.00062	.45	.00043	.00615	.00453	.00176	.00035	.0084
%RSD	19.542	.39163	21.028	5.3725	80.076	549.83	10.037	.16741
#1	.00290	113.69	.00211	.12072	.00935	.00043	.00314	5.0497
#2	.00278	114.20	.00240	.11417	.00060	.00094	.00356	5.0330
#3	.00392	114.58	.00156	.10843	.00702	-.00233	.00385	5.0397
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115001 Acquired: 6/25/2016 14:23:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00003**	**.24066**	**.01417**	**-.00245**	**.00766**	**.00892**	**.00404**
Stddev	.00021	.00154	.00348	.00225	.00065	.00011	.00127
%RSD	805.05	.63820	24.568	91.853	8.5297	1.1926	31.485
#1	.00024	.23889	.01059	-.00116	.00840	.00901	.00275
#2	.00001	.24171	.01754	-.00114	.00737	.00895	.00530
#3	-.00018	.24137	.01438	-.00504	.00719	.00880	.00407
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9688.5**	**58932.**	**3057.0**
Stddev	25.8	93.	13.7
%RSD	.26652	.15746	.44955
#1	9698.0	58835.	3047.2
#2	9708.3	59020.	3072.7
#3	9659.3	58940.	3051.0

Approved: June 27, 2016

Sample Name: L1606115002 Acquired: 6/25/2016 14:27:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00203	.48878	.00167	.06520	.04585	.00006	17.152	.00001
Stddev	.00268	.01554	.00141	.00090	.00204	.00009	.173	.00009
%RSD	131.70	3.1791	84.852	1.3869	4.4598	167.56	1.0064	934.85
#1	.00187	.50660	.00286	.06544	.04801	.00015	17.256	-.00009
#2	-.00056	.47801	.00204	.06596	.04394	-.00004	16.953	.00007
#3	.00479	.48175	.00010	.06420	.04559	.00005	17.247	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00031	.00168	.00243	.39846	4.7446	.00319	5.4375	.01729
Stddev	.00031	.00066	.00184	.00667	.1185	.00289	.0385	.00153
%RSD	100.89	39.152	75.822	1.6749	2.4966	90.823	.70870	8.8325
#1	.00065	.00096	.00122	.39954	4.7392	.00001	5.4553	.01884
#2	.00006	.00183	.00151	.39131	4.6289	.00567	5.3932	.01727
#3	.00020	.00224	.00454	.40452	4.8657	.00387	5.4638	.01578
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00028	38.880	.00010	.07228	.00433	-.00020	.00027	1.5503
Stddev	.00033	.402	.00056	.00820	.00375	.00324	.00590	.0251
%RSD	119.49	1.0347	568.40	11.341	86.648	1601.4	2189.5	1.6195
#1	-.00010	39.210	-.00001	.06867	.00198	.00127	.00262	1.5658
#2	.00053	38.432	.00070	.06650	.00235	-.00392	-.00644	1.5637
#3	.00040	38.999	-.00040	.08166	.00865	.00205	.00464	1.5213
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115002 Acquired: 6/25/2016 14:27:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00004	.09712	.00701	-.00122	.00599	.00508	.00239
Stddev	.00062	.00057	.00249	.00352	.00080	.00018	.00169
%RSD	1544.5	.58817	35.547	288.53	13.353	3.5369	70.760
#1	.00058	.09776	.00445	-.00524	.00624	.00503	.00322
#2	-.00064	.09666	.00715	.00026	.00510	.00528	.00044
#3	.00018	.09693	.00942	.00132	.00664	.00493	.00349
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9845.3	59884.	3057.5
Stddev	61.7	141.	28.2
%RSD	.62651	.23463	.92251
#1	9792.5	59732.	3034.3
#2	9830.2	60008.	3088.9
#3	9913.1	59914.	3049.2

Approved: June 27, 2016

Sample Name: L1606115102 Acquired: 6/25/2016 14:31:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00159	.14924	.00211	.10945	.07344	.00012	35.265	.00011
Stddev	.00084	.00304	.00150	.00214	.00141	.00005	.199	.00019
%RSD	53.011	2.0340	71.147	1.9507	1.9218	40.963	.56552	183.34
#1	.00069	.14626	.00205	.10856	.07211	.00007	35.174	.00027
#2	.00171	.15233	.00364	.10790	.07329	.00017	35.493	-.00011
#3	.00236	.14912	.00064	.11188	.07492	.00013	35.127	.00016
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00052	.00053	.00363	.24070	2.2004	.00754	18.847	.05800
Stddev	.00037	.00064	.00076	.00439	.0134	.00192	.227	.00368
%RSD	71.691	119.20	20.837	1.8247	.61086	25.451	1.2033	6.3535
#1	.00009	-.00001	.00323	.23564	2.1949	.00870	18.615	.05962
#2	.00071	.00038	.00450	.24286	2.2158	.00532	19.068	.05378
#3	.00075	.00124	.00316	.24359	2.1907	.00860	18.858	.06059
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00267	93.416	.00066	.05411	.00356	-.00299	.00235	4.9794
Stddev	.00025	.232	.00043	.00846	.00271	.00266	.00452	.0235
%RSD	9.2843	.24821	65.257	15.640	76.210	88.939	192.50	.47215
#1	.00239	93.189	.00082	.06384	.00043	-.00008	-.00245	4.9968
#2	.00274	93.652	.00099	.04849	.00522	-.00529	.00653	4.9888
#3	.00288	93.408	.00017	.05000	.00501	-.00360	.00296	4.9527
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115102 Acquired: 6/25/2016 14:31:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00015	.22720	-.00049	-.00434	.00477	.01199	.00282
Stddev	.00094	.00191	.00341	.00316	.00052	.00018	.00083
%RSD	624.90	.83946	689.70	72.807	10.840	1.4772	29.345
#1	.00091	.22526	.00103	-.00351	.00496	.01201	.00271
#2	-.00090	.22907	-.00440	-.00783	.00516	.01216	.00370
#3	.00043	.22727	.00189	-.00168	.00418	.01181	.00206
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9596.9	58682.	3041.4
Stddev	35.1	192.	10.5
%RSD	.36599	.32727	.34598
#1	9558.5	58834.	3032.3
#2	9605.0	58747.	3039.1
#3	9627.4	58466.	3052.9

Approved: June 27, 2016

Sample Name: L1606115105 Acquired: 6/25/2016 14:35:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00046	.08611	.00385	.13923	.09101	.00009	44.089	.00002
Stddev	.00149	.00500	.00146	.00310	.00028	.00009	.047	.00004
%RSD	320.61	5.8036	37.928	2.2254	.30236	102.05	.10587	146.93
#1	-.00016	.08678	.00262	.14146	.09069	.00015	44.139	.00006
#2	-.00061	.09074	.00347	.13569	.09114	.00014	44.082	.00002
#3	.00216	.08081	.00547	.14055	.09119	-.00002	44.046	-.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	.00077	.00300	.22257	1.9197	.00784	24.665	.07651
Stddev	.00017	.00044	.00144	.01624	.0794	.00414	.178	.00017
%RSD	34.605	57.594	48.079	7.2986	4.1341	52.815	.72144	.22200
#1	.00066	.00116	.00292	.22095	1.8333	.01224	24.592	.07669
#2	.00046	.00029	.00448	.23956	1.9894	.00402	24.534	.07635
#3	.00033	.00086	.00160	.20719	1.9365	.00726	24.867	.07649
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00251	121.52	.00039	.04724	.00176	-.00202	.00114	5.9103
Stddev	.00040	.42	.00024	.00471	.00159	.00145	.00478	.0521
%RSD	16.119	.34732	62.165	9.9710	90.693	71.551	420.31	.88218
#1	.00228	121.85	.00064	.05267	.00005	-.00053	.00361	5.8778
#2	.00297	121.67	.00016	.04481	.00202	-.00342	.00418	5.9704
#3	.00227	121.04	.00037	.04425	.00320	-.00211	-.00437	5.8826
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115105 Acquired: 6/25/2016 14:35:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00058	.28912	-.00001	-.00113	.00539	.00821	.00159
Stddev	.00104	.00030	.00390	.00110	.00033	.00010	.00078
%RSD	180.18	.10436	27479.	97.280	6.0662	1.2671	48.813
#1	.00169	.28878	.00322	.00014	.00577	.00817	.00171
#2	.00041	.28919	.00108	-.00182	.00517	.00833	.00230
#3	-.00037	.28937	-.00434	-.00172	.00524	.00813	.00076
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9591.1	58078.	3001.7
Stddev	50.0	111.	7.0
%RSD	.52084	.19101	.23475
#1	9648.1	57963.	2993.9
#2	9554.9	58185.	3007.6
#3	9570.3	58085.	3003.7

Approved: June 27, 2016

Sample Name: L1606115401 Acquired: 6/25/2016 14:39:13 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00028**	**.06369**	**.00211**	**1.6713**	**.01926**	**.00010**	F **452.03**
Stddev	.00322	.00656	.00351	.0005	.00118	.00010	.89
%RSD	1157.9	10.295	166.24	.02810	6.1016	104.76	.19751
#1	.00114	.06870	-.00193	1.6717	.01795	.00008	451.94
#2	.00198	.05626	.00441	1.6715	.02023	.00000	452.97
#3	-.00396	.06610	.00386	1.6708	.01959	.00021	451.19
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail
High Limit							270.00
Low Limit							-.10000

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00042**	**.00175**	**.00336**	**.00358**	**.08657**	**4.0470**	**.10636**
Stddev	.00024	.00006	.00098	.00171	.01546	.0608	.00217
%RSD	56.333	3.4162	29.186	47.723	17.854	1.5017	2.0382
#1	.00015	.00180	.00230	.00410	.09591	4.0001	.10764
#2	.00053	.00168	.00424	.00167	.09508	4.1156	.10386
#3	.00059	.00177	.00354	.00498	.06873	4.0251	.10759
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**104.45**	**.25861**	**.02316**	F **889.97**	**.00046**	**.03619**	**.00320**
Stddev	.32	.00175	.00063	8.94	.00094	.00407	.00206
%RSD	.30573	.67823	2.7369	1.0045	203.65	11.246	64.428
#1	104.40	.26014	.02389	897.40	-.00019	.03795	.00082
#2	104.79	.25898	.02276	892.47	.00155	.03153	.00442
#3	104.16	.25670	.02283	880.05	.00003	.03907	.00435
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 27, 2016

Sample Name: L1606115401 Acquired: 6/25/2016 14:39:13 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-02

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**.04556**	**5.0652**	**-.00092**	**8.3910**	F **-.05563**	**-.00220**
Stddev	.00340	.00461	.0400	.00030	.0291	.00259	.00257
%RSD	2618.0	10.119	.78890	32.506	.34723	4.6498	116.61
#1	-.00398	.04208	5.1054	-.00114	8.4061	-.05807	-.00507
#2	.00246	.05079	5.0645	-.00058	8.4095	-.05590	-.00010
#3	.00113	.04380	5.0255	-.00104	8.3574	-.05292	-.00145
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.01649**	**.00488**	**.00303**
Stddev	.00056	.00014	.00088
%RSD	3.4018	2.9544	29.046
#1	.01604	.00505	.00374
#2	.01712	.00480	.00205
#3	.01632	.00480	.00329
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**8680.3**	**53198.**	**2985.6**
Stddev	180.6	1315.	22.8
%RSD	2.0805	2.4712	.76305
#1	8472.3	54211.	3004.6
#2	8770.9	53671.	2960.4
#3	8797.6	51712.	2991.9

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 14:43:13 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.41555	10.459	.41251	.50877	1.0374	.05191	10.456
Stddev	.00198	.038	.00049	.00412	.0216	.00011	.214
%RSD	.47730	.36575	.11809	.80911	2.0778	.21811	2.0454
#1	.41688	10.435	.41304	.50472	1.0594	.05178	10.677
#2	.41327	10.438	.41238	.51295	1.0163	.05199	10.250
#3	.41650	10.503	.41210	.50863	1.0364	.05195	10.439
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05257	.20726	.51590	.53004	4.1313	50.651	1.0395
Stddev	.00009	.00077	.00185	.00200	.0874	1.016	.0162
%RSD	.16567	.37212	.35803	.37798	2.1147	2.0061	1.5578
#1	.05256	.20724	.51397	.52980	4.1823	51.735	1.0578
#2	.05249	.20650	.51606	.52817	4.0304	49.720	1.0270
#3	.05266	.20804	.51766	.53216	4.1811	50.499	1.0337
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	10.150	.52278	1.0570	51.369	.52494	10.305	.52104
Stddev	.208	.00621	.0043	.990	.00089	.023	.00138
%RSD	2.0459	1.1888	.41022	1.9281	.16882	.22613	.26480
#1	10.382	.52905	1.0620	52.419	.52481	10.311	.51949
#2	10.083	.51662	1.0546	50.452	.52412	10.279	.52151
#3	9.9835	.52267	1.0544	51.234	.52588	10.325	.52213
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 14:43:13 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2435**	**.41340**	**5.3297**	**1.0365**	**1.0404**	**1.0490**	**.52867**
Stddev	.0031	.00239	.0078	.0027	.0194	.0184	.00366
%RSD	.25127	.57838	.14640	.26182	1.8676	1.7498	.69286
#1	1.2461	.41607	5.3340	1.0379	1.0597	1.0678	.52786
#2	1.2400	.41146	5.3207	1.0334	1.0208	1.0311	.53267
#3	1.2443	.41266	5.3345	1.0383	1.0408	1.0483	.52548
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0507**	**1.0444**	F **3.0477**
Stddev	.0007	.0005	.0535
%RSD	.06848	.05207	1.7552
#1	1.0503	1.0446	3.1080
#2	1.0502	1.0438	3.0059
#3	1.0515	1.0449	3.0292
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9539.3**	**57751.**	**2940.1**
Stddev	31.6	140.	48.7
%RSD	.33120	.24295	1.6549
#1	9524.0	57816.	2886.9
#2	9575.6	57847.	2982.3
#3	9518.3	57590.	2951.2

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 14:46:39 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00165**	**.00284**	**.00176**	**.00388**	**.00113**	**.00012**	**-.05362**	**.00014**
Stddev	.00125	.01178	.00236	.00154	.00084	.00005	.04538	.00017
%RSD	75.692	415.29	134.23	39.663	74.227	41.354	84.617	118.22
#1	.00309	-.01077	.00340	.00450	.00053	.00007	-.00140	.00006
#2	.00089	.00942	.00281	.00212	.00078	.00011	-.07612	.00034
#3	.00097	.00986	-.00094	.00500	.00209	.00017	-.08336	.00003
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00024**	**.00081**	**-.00057**	**.03333**	**.05011**	**.00030**	**.06903**	**-.00046**
Stddev	.00028	.00137	.00241	.00857	.04565	.00118	.08939	.00080
%RSD	113.95	168.04	420.16	25.719	91.103	392.02	129.50	172.20
#1	.00053	-.00076	.00045	.02973	.09790	.00058	.12386	.00033
#2	.00023	.00164	-.00333	.02715	.04546	-.00100	-.03412	-.00045
#3	-.00003	.00157	.00116	.04312	.00696	.00132	.11734	-.00127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00329**	**.09636**	**-.00015**	**.00820**	**-.00027**	**.00540**	**.00512**	**-.00542**
Stddev	.00036	.02670	.00052	.00425	.00262	.00178	.00316	.00074
%RSD	10.874	27.706	348.06	51.818	968.89	32.864	61.650	13.619
#1	.00316	.11195	.00032	.01302	.00232	.00377	.00634	-.00562
#2	.00301	.11159	-.00071	.00662	-.00021	.00729	.00749	-.00604
#3	.00369	.06553	-.00007	.00498	-.00292	.00514	.00154	-.00460
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 14:46:39 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00030	.00084	-.00314	-.00152	.00124	-.00002	.00422
Stddev	.00043	.00035	.00315	.00163	.00162	.00008	.00071
%RSD	142.14	41.948	100.51	107.27	131.28	387.02	16.743
#1	.00054	.00124	-.00369	-.00248	.00007	.00002	.00481
#2	.00057	.00059	.00026	.00036	.00055	-.00012	.00343
#3	-.00019	.00069	-.00597	-.00244	.00309	.00004	.00441
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9509.0	57769.	2936.8
Stddev	19.9	100.	17.6
%RSD	.20940	.17266	.59912
#1	9496.2	57883.	2947.5
#2	9532.0	57728.	2946.4
#3	9498.9	57697.	2916.5

Approved: June 27, 2016

Sample Name: L1606115402S Acquired: 6/25/2016 14:50:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.20060**	**4.9650**	**.19517**	**2.5768**	**.50183**	**.02406**	F **444.38**
Stddev	.00064	.0130	.00068	.0061	.00191	.00013	.71
%RSD	.32075	.26173	.35003	.23805	.38088	.54060	.15891
#1	.20131	4.9565	.19473	2.5781	.50380	.02409	444.63
#2	.20005	4.9586	.19596	2.5701	.50170	.02392	443.58
#3	.20045	4.9800	.19484	2.5821	.49998	.02418	444.92
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail
High Limit							270.00
Low Limit							-.10000

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02442**	**.09295**	**.23551**	**.23263**	**1.9191**	**28.985**	**.61166**
Stddev	.00012	.00006	.00093	.00115	.0123	.093	.00400
%RSD	.49172	.06109	.39479	.49604	.64319	.32038	.65326
#1	.02431	.09291	.23631	.23296	1.9231	29.086	.60738
#2	.02455	.09302	.23573	.23359	1.9290	28.966	.61529
#3	.02440	.09293	.23449	.23135	1.9053	28.904	.61230
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**105.95**	**.47454**	**.51043**	F **884.75**	**.22811**	**4.9353**	**.22750**
Stddev	.25	.00567	.00309	9.80	.00122	.0272	.00524
%RSD	.23467	1.1953	.60519	1.1073	.53281	.55168	2.3044
#1	105.68	.47627	.51271	882.43	.22677	4.9309	.23221
#2	106.17	.47914	.51168	876.32	.22915	4.9645	.22843
#3	105.99	.46820	.50692	895.49	.22841	4.9105	.22185
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 27, 2016

Sample Name: L1606115402S Acquired: 6/25/2016 14:50:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-07

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.58292**	**.23516**	**7.5230**	**.45764**	**8.4638**	**.43138**	**.22024**
Stddev	.00661	.00590	.0386	.00329	.0240	.00658	.00196
%RSD	1.1347	2.5090	.51250	.71986	.28403	1.5254	.88967
#1	.58964	.23313	7.5444	.46024	8.4871	.43835	.22044
#2	.58270	.23054	7.5461	.45874	8.4651	.43053	.22209
#3	.57642	.24181	7.4785	.45394	8.4391	.42527	.21819
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.50974**	**.45745**	**.02017**
Stddev	.00182	.00152	.00279
%RSD	.35750	.33182	13.807
#1	.50798	.45815	.01724
#2	.50964	.45849	.02049
#3	.51162	.45571	.02279
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**8784.0**	**53813.**	**2939.5**
Stddev	58.9	564.	26.0
%RSD	.67031	1.0483	.88591
#1	8819.5	54202.	2951.8
#2	8816.5	54070.	2957.2
#3	8716.1	53166.	2909.6

Approved: June 27, 2016

Sample Name: L1606115403SD Acquired: 6/25/2016 14:54:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19884**	**5.0097**	**.20311**	**2.5960**	**.50146**	**.02443**	F **447.43**
Stddev	.00024	.0049	.00234	.0007	.00093	.00023	1.15
%RSD	.12160	.09827	1.1515	.02603	.18621	.94697	.25628
#1	.19893	5.0097	.20547	2.5958	.50254	.02464	447.39
#2	.19857	5.0047	.20308	2.5967	.50101	.02419	446.31
#3	.19903	5.0146	.20079	2.5954	.50084	.02448	448.60
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail
High Limit							270.00
Low Limit							-.10000

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02483**	**.09451**	**.23821**	**.23915**	**1.9535**	**28.902**	**.60873**
Stddev	.00032	.00036	.00123	.00189	.0137	.108	.00282
%RSD	1.3026	.38369	.51502	.78901	.69899	.37351	.46252
#1	.02509	.09492	.23941	.23914	1.9479	29.020	.60854
#2	.02447	.09421	.23696	.23726	1.9691	28.808	.61163
#3	.02493	.09441	.23826	.24103	1.9435	28.877	.60601
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**106.60**	**.48110**	**.52231**	F **881.76**	**.23168**	**5.0586**	**.22685**
Stddev	.66	.00483	.00325	3.70	.00136	.0426	.00410
%RSD	.61507	1.0048	.62269	.41950	.58652	.84170	1.8079
#1	107.35	.47570	.52433	877.70	.23164	5.0964	.23130
#2	106.14	.48502	.51856	884.94	.23034	5.0125	.22604
#3	106.32	.48257	.52404	882.64	.23305	5.0670	.22322
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 27, 2016

Sample Name: L1606115403SD Acquired: 6/25/2016 14:54:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573707-08

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.59871**	**.24615**	**7.5687**	**.46699**	**8.5686**	**.42893**	**.22250**
Stddev	.00609	.00268	.0516	.00318	.0150	.00524	.00079
%RSD	1.0180	1.0870	.68153	.68125	.17550	1.2224	.35602
#1	.60575	.24920	7.6059	.46994	8.5810	.42717	.22334
#2	.59510	.24505	7.5098	.46362	8.5519	.42478	.22177
#3	.59529	.24419	7.5903	.46740	8.5730	.43482	.22237
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.51554**	**.46791**	**.00590**
Stddev	.00333	.00294	.00185
%RSD	.64552	.62860	31.408
#1	.51938	.47002	.00382
#2	.51356	.46455	.00651
#3	.51367	.46915	.00737
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**8900.3**	**52138.**	**2849.7**
Stddev	61.9	226.	11.1
%RSD	.69538	.43440	.38823
#1	8858.0	51879.	2838.3
#2	8971.3	52295.	2860.4
#3	8871.6	52242.	2850.4

Approved: June 27, 2016

Sample Name: L1606115404 Acquired: 6/25/2016 14:58:09 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.00639	-.00036	.01270	.00241	.00013	.28017	.00009
Stddev	.00146	.00656	.00244	.00114	.00043	.00008	.03044	.00010
%RSD	248.07	102.69	686.61	9.0101	18.005	61.828	10.864	104.11
#1	.00089	-.00046	-.00313	.01156	.00227	.00022	.24651	.00012
#2	.00188	.01261	.00061	.01385	.00206	.00012	.28824	-.00001
#3	-.00100	.00700	.00146	.01270	.00289	.00005	.30576	.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	.01305	.00240	.07792	.04064	-.00228	.12469	.00253
Stddev	.00039	.00070	.00120	.00694	.09353	.00295	.18212	.00145
%RSD	747.32	5.3526	50.113	8.9064	230.14	129.38	146.06	57.512
#1	.00045	.01376	.00301	.08066	.00523	-.00157	.14106	.00090
#2	-.00033	.01305	.00317	.08306	.14671	-.00552	.29807	.00300
#3	.00004	.01236	.00101	.07002	-.03002	.00025	-.06507	.00369
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00228	.75291	.00807	.00690	.00300	-.00255	.00146	.13135
Stddev	.00024	.01638	.00065	.00315	.00148	.00424	.00328	.00151
%RSD	10.641	2.1753	8.1061	45.563	49.471	166.69	225.01	1.1485
#1	.00245	.73694	.00868	.00525	.00160	-.00242	-.00127	.13296
#2	.00238	.75210	.00738	.01053	.00455	-.00685	.00055	.13112
#3	.00200	.76967	.00814	.00493	.00285	.00163	.00509	.12997
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115404 Acquired: 6/25/2016 14:58:09 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00160	.00282	.00431	-.00324	.00105	.00371	.00033
Stddev	.00036	.00041	.00156	.00193	.00089	.00012	.00066
%RSD	22.302	14.434	36.136	59.618	84.281	3.3596	201.18
#1	.00136	.00295	.00420	-.00210	.00202	.00385	-.00028
#2	.00201	.00236	.00592	-.00547	.00085	.00361	.00023
#3	.00143	.00314	.00281	-.00214	.00028	.00369	.00104
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9713.7	59908.	3002.7
Stddev	32.6	164.	34.0
%RSD	.33561	.27353	1.1337
#1	9740.1	60011.	2965.4
#2	9677.2	59995.	3032.2
#3	9723.7	59719.	3010.3

Approved: June 27, 2016

Sample Name: L1606115501 Acquired: 6/25/2016 15:02:07 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00226	.00780	.00401	.26646	.17918	.00000	45.474	-.00004
Stddev	.00287	.00301	.00103	.00133	.00076	.00004	.222	.00008
%RSD	126.68	38.602	25.778	.50014	.42327	919.30	.48843	228.57
#1	.00509	.01128	.00462	.26750	.17974	.00002	45.718	.00006
#2	.00233	.00600	.00459	.26694	.17949	-.00004	45.422	-.00008
#3	-.00064	.00613	.00281	.26496	.17832	.00004	45.283	-.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.00088	.00134	.05774	1.5654	.00884	27.038	1.6699
Stddev	.00027	.00028	.00194	.01943	.1007	.00205	.083	.0027
%RSD	44.880	31.682	144.85	33.650	6.4312	23.214	.30796	.16065
#1	.00083	.00056	-.00088	.03531	1.4495	.00687	26.967	1.6730
#2	.00063	.00109	.00219	.06938	1.6158	.01096	27.130	1.6684
#3	.00031	.00099	.00272	.06852	1.6309	.00869	27.018	1.6684
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00414	74.521	.00232	.02280	.00145	-.00333	.00418	6.9589
Stddev	.00012	.362	.00063	.00622	.00648	.00099	.00335	.0172
%RSD	2.8693	.48589	27.089	27.292	447.38	29.598	80.140	.24675
#1	.00419	74.939	.00163	.02521	-.00229	-.00322	.00046	6.9755
#2	.00400	74.314	.00245	.02746	-.00229	-.00437	.00510	6.9601
#3	.00421	74.311	.00286	.01573	.00893	-.00240	.00696	6.9412
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115501 Acquired: 6/25/2016 15:02:07 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00017**	**.31080**	**-.00408**	**-.00246**	**.00074**	**.00517**	**.00097**
Stddev	.00101	.00127	.01220	.00176	.00032	.00009	.00026
%RSD	589.50	.40720	299.39	71.557	43.195	1.7947	26.599
#1	.00067	.31223	.00838	-.00158	.00040	.00528	.00122
#2	-.00129	.30982	-.01601	-.00449	.00079	.00510	.00071
#3	.00011	.31036	-.00459	-.00131	.00103	.00514	.00097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9554.2**	**58064.**	**2987.7**
Stddev	24.6	152.	17.9
%RSD	.25722	.26211	.59940
#1	9527.5	57892.	2971.4
#2	9576.0	58182.	2984.9
#3	9559.0	58118.	3006.9

Approved: June 27, 2016

Sample Name: L1606115502 Acquired: 6/25/2016 15:05:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00077	.00104	.00250	.19417	.33327	.00001	49.969	.00011
Stddev	.00225	.00260	.00021	.00180	.00208	.00007	.144	.00004
%RSD	293.51	250.20	8.2634	.92950	.62356	535.11	.28842	36.458
#1	.00053	-.00100	.00271	.19484	.33088	.00006	50.109	.00015
#2	-.00336	.00396	.00251	.19555	.33470	.00004	49.821	.00012
#3	.00053	.00015	.00229	.19213	.33422	-.00007	49.978	.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00029	.00039	.00161	4.4932	.30074	.00699	33.094	.46968
Stddev	.00024	.00046	.00055	.0210	.07361	.00374	.139	.00201
%RSD	82.593	118.92	34.456	.46819	24.475	53.562	.42105	.42764
#1	.00002	-.00012	.00218	4.4716	.35097	.00318	33.254	.46985
#2	.00046	.00051	.00108	4.5136	.21625	.00712	33.021	.47160
#3	.00039	.00078	.00156	4.4944	.33501	.01067	33.006	.46759
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00134	56.528	.00048	.02851	.00540	-.00461	.00421	7.9465
Stddev	.00013	.068	.00114	.00462	.00195	.00412	.00583	.0128
%RSD	9.8009	.11946	234.80	16.220	36.120	89.402	138.37	.16106
#1	.00123	56.559	.00179	.03249	.00375	-.00852	-.00246	7.9518
#2	.00130	56.451	-.00006	.02962	.00489	-.00500	.00681	7.9559
#3	.00149	56.575	-.00027	.02344	.00755	-.00031	.00829	7.9320
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115502 Acquired: 6/25/2016 15:05:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00013**	**.41936**	**-.00337**	**-.00247**	**.00035**	**.00337**	**.00091**
Stddev	.00022	.00050	.00295	.00045	.00041	.00006	.00160
%RSD	168.21	.11997	87.675	18.352	119.11	1.8649	175.07
#1	.00016	.41971	-.00131	-.00286	-.00012	.00341	-.00093
#2	.00033	.41958	-.00675	-.00259	.00050	.00330	.00173
#3	-.00010	.41878	-.00204	-.00197	.00066	.00341	.00194
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9627.3**	**58565.**	**2988.6**
Stddev	17.2	39.	5.1
%RSD	.17881	.06667	.17070
#1	9625.0	58521.	2993.1
#2	9645.6	58579.	2989.5
#3	9611.4	58596.	2983.0

Approved: June 27, 2016

Sample Name: L1606115503 Acquired: 6/25/2016 15:09:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00133	.01981	.00070	.15980	.12508	.00001	47.416	.00012
Stddev	.00149	.00690	.00351	.00313	.00043	.00010	.201	.00026
%RSD	112.43	34.815	504.28	1.9589	.34619	838.85	.42403	214.19
#1	.00108	.02777	-.00104	.15681	.12534	.00005	47.373	.00032
#2	-.00003	.01583	-.00161	.16306	.12532	.00009	47.635	.00022
#3	.00292	.01581	.00474	.15953	.12458	-.00010	47.240	-.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00011	.00032	.00351	1.3413	.87483	.00900	17.381	.47700
Stddev	.00046	.00045	.00082	.0140	.08308	.00172	.089	.00081
%RSD	429.81	139.46	23.247	1.0398	9.4971	19.124	.51021	.16934
#1	.00045	.00052	.00286	1.3297	.89461	.00799	17.319	.47725
#2	-.00041	.00065	.00443	1.3568	.78364	.01099	17.483	.47609
#3	.00028	-.00019	.00325	1.3373	.94624	.00803	17.342	.47764
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00026	48.369	.00082	.10919	.00167	-.00062	.00179	5.8799
Stddev	.00025	.203	.00140	.00660	.00170	.00395	.00827	.0280
%RSD	97.415	.41932	171.45	6.0459	101.94	641.48	462.74	.47549
#1	-.00003	48.343	.00233	.10833	.00101	-.00514	.00952	5.8933
#2	-.00053	48.584	.00055	.10306	.00039	.00119	.00277	5.8478
#3	-.00022	48.181	-.00043	.11618	.00360	.00211	-.00693	5.8986
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115503 Acquired: 6/25/2016 15:09:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00019**	**.26600**	**-.00537**	**-.00338**	**.00017**	**.00914**	**.00020**
Stddev	.00067	.00095	.00241	.00310	.00116	.00022	.00114
%RSD	345.33	.35889	44.987	91.459	671.24	2.4286	566.35
#1	-.00057	.26556	-.00315	-.00551	-.00106	.00889	.00099
#2	.00050	.26710	-.00794	.00017	.00033	.00931	-.00111
#3	.00065	.26535	-.00501	-.00481	.00125	.00922	.00072
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9587.2**	**58263.**	**2972.1**
Stddev	19.6	279.	4.9
%RSD	.20409	.47925	.16604
#1	9598.2	58004.	2969.9
#2	9598.7	58226.	2968.6
#3	9564.6	58559.	2977.8

Approved: June 27, 2016

Sample Name: L1606117801 Acquired: 6/25/2016 15:13:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00123	.01562	.00788	.04060	.02600	.00012	2.4592	.00007
Stddev	.00244	.00616	.00118	.00330	.00129	.00005	.0606	.00006
%RSD	197.99	39.461	15.044	8.1358	4.9545	44.901	2.4620	79.703
#1	.00387	.02096	.00725	.04075	.02646	.00010	2.5149	.00002
#2	-.00097	.00887	.00713	.04382	.02700	.00008	2.4680	.00013
#3	.00081	.01703	.00924	.03722	.02455	.00018	2.3948	.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00003	.00074	.00609	.02685	.43656	.00432	.57708	-.00111
Stddev	.00033	.00144	.00018	.01358	.05030	.00166	.05180	.00137
%RSD	1012.2	194.14	2.9070	50.575	11.522	38.411	8.9754	123.48
#1	-.00025	.00070	.00629	.04250	.40355	.00528	.52291	-.00216
#2	.00039	.00219	.00602	.01809	.49445	.00241	.62611	-.00159
#3	-.00004	-.00068	.00595	.01997	.41167	.00528	.58223	.00044
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00299	95.163	-.00070	.01050	.00222	-.00635	.00641	3.4783
Stddev	.00029	.186	.00058	.00517	.00074	.00345	.00307	.0048
%RSD	9.6245	.19542	82.717	49.206	33.276	54.334	47.947	.13892
#1	.00325	95.230	-.00082	.01297	.00142	-.01032	.00356	3.4792
#2	.00268	95.306	-.00121	.00456	.00236	-.00408	.00600	3.4827
#3	.00303	94.953	-.00007	.01396	.00287	-.00464	.00966	3.4731
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606117801 Acquired: 6/25/2016 15:13:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00037	.11699	.00044	-.00188	.00105	.00210	.00054
Stddev	.00103	.00010	.00531	.00109	.00009	.00007	.00166
%RSD	275.08	.08958	1212.8	57.943	8.6567	3.2366	310.87
#1	.00154	.11703	-.00019	-.00063	.00114	.00212	-.00081
#2	-.00042	.11707	-.00453	-.00235	.00105	.00215	.00002
#3	-.00000	.11687	.00604	-.00264	.00096	.00202	.00240
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9648.4	58829.	2994.3
Stddev	5.2	37.	6.7
%RSD	.05359	.06360	.22320
#1	9653.7	58809.	2992.7
#2	9643.3	58806.	2988.6
#3	9648.3	58873.	3001.7

Approved: June 27, 2016

Sample Name: L1606122701 Acquired: 6/25/2016 15:17:43 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00116	.09783	-.00055	.00242	.01519	.00006	7.7326	-.00002
Stddev	.00092	.00377	.00239	.00147	.00198	.00002	.1171	.00010
%RSD	79.355	3.8537	437.80	60.740	13.003	38.615	1.5138	459.39
#1	-.00047	.10163	-.00294	.00351	.01401	.00007	7.7569	.00006
#2	-.00220	.09409	.00184	.00075	.01747	.00007	7.8356	-.00013
#3	-.00081	.09776	-.00054	.00300	.01410	.00003	7.6053	.00000
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00027	.00227	.00167	.03463	.33602	-.00038	-.01109	-.00127
Stddev	.00016	.00059	.00051	.00142	.06178	.00154	.10836	.00134
%RSD	58.524	25.933	30.520	4.0976	18.384	404.83	977.09	105.32
#1	.00014	.00159	.00133	.03617	.30512	-.00005	-.09610	-.00256
#2	.00023	.00261	.00225	.03337	.29580	-.00206	.11092	-.00136
#3	.00045	.00260	.00142	.03436	.40715	.00097	-.04809	.00011
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00015	2.3350	.00044	.00731	.00430	-.00426	.00383	1.1328
Stddev	.00020	.0307	.00076	.00233	.00100	.00159	.00576	.0032
%RSD	132.69	1.3153	170.74	31.908	23.273	37.398	150.44	.27905
#1	.00005	2.3000	.00116	.00516	.00531	-.00538	.00306	1.1309
#2	-.00036	2.3574	.00052	.00700	.00430	-.00496	-.00151	1.1310
#3	-.00016	2.3476	-.00035	.00979	.00330	-.00244	.00992	1.1364
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606122701 Acquired: 6/25/2016 15:17:43 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	.25125	.00414	.00180	.00364	.00056	.00136
Stddev	.00018	.00103	.01324	.00145	.00074	.00007	.00130
%RSD	364.81	.41002	320.05	80.159	20.271	13.218	96.223
#1	.00011	.25131	-.01092	.00207	.00450	.00061	.00115
#2	.00019	.25224	.00936	.00024	.00328	.00047	.00016
#3	-.00016	.25018	.01397	.00310	.00316	.00058	.00275
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9969.0	60494.	3048.9
Stddev	15.1	129.	13.3
%RSD	.15159	.21283	.43564
#1	9969.3	60641.	3055.8
#2	9983.9	60438.	3033.6
#3	9953.7	60402.	3057.3

Approved: June 27, 2016

Sample Name: L1606122706 Acquired: 6/25/2016 15:21:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00063**	**.00567**	**.00083**	**.00123**	**.02806**	**.00015**	**13.472**	**-.00038**
Stddev	.00250	.00554	.00101	.00202	.00007	.00007	.038	.00019
%RSD	395.00	97.673	121.57	163.43	.25882	47.682	.28032	49.732
#1	.00247	-.00071	.00042	-.00047	.02810	.00007	13.487	-.00023
#2	.00165	.00924	.00199	.00346	.02810	.00021	13.429	-.00032
#3	-.00222	.00849	.00010	.00072	.02797	.00018	13.501	-.00060
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00101**	**.00061**	**.00157**	**.10367**	**.20316**	**.00472**	**8.1185**	**.10636**
Stddev	.00008	.00062	.00064	.01594	.00836	.00156	.1274	.00158
%RSD	8.0012	102.28	40.873	15.374	4.1155	32.981	1.5689	1.4897
#1	.00092	-.00009	.00190	.11049	.21157	.00610	7.9808	.10550
#2	.00107	.00111	.00198	.11507	.19485	.00503	8.2322	.10539
#3	.00104	.00079	.00083	.08546	.20306	.00303	8.1425	.10819
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00044**	**33.974**	**.00004**	**.00429**	**-.00079**	**-.00333**	**.00187**	**.74797**
Stddev	.00043	.239	.00011	.00671	.00165	.00196	.00470	.01610
%RSD	97.916	.70374	251.40	156.52	208.37	58.883	251.58	2.1527
#1	-.00064	33.727	-.00007	.01191	-.00018	-.00196	-.00208	.73053
#2	-.00073	33.993	.00015	-.00074	.00047	-.00558	.00061	.75112
#3	.00005	34.204	.00006	.00169	-.00266	-.00246	.00707	.76226
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606122706 Acquired: 6/25/2016 15:21:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00014**	**.30857**	**.00233**	**-.00259**	**.00019**	**.00158**	**.00089**
Stddev	.00012	.00270	.00311	.00390	.00060	.00013	.00062
%RSD	87.894	.87470	133.51	150.50	319.58	8.2216	69.559
#1	-.00001	.30602	.00364	-.00319	-.00046	.00143	.00122
#2	-.00015	.30828	-.00122	.00157	.00073	.00163	.00127
#3	-.00026	.31139	.00457	-.00615	.00030	.00167	.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9924.8**	**59991.**	**2995.5**
Stddev	18.6	172.	10.6
%RSD	.18782	.28732	.35313
#1	9924.0	59808.	2993.5
#2	9943.8	60150.	3007.0
#3	9906.6	60016.	2986.1

Approved: June 27, 2016

Sample Name: L1606127902 Acquired: 6/25/2016 15:25:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00298**	**.77289**	**.00033**	**.00803**	**.01053**	**.00016**	**9.3649**	**-.00023**
Stddev	.00188	.01227	.00104	.00267	.00091	.00007	.0910	.00007
%RSD	63.015	1.5872	311.80	33.213	8.6722	42.111	.97197	32.230
#1	.00105	.75975	.00114	.01109	.01057	.00009	9.3786	-.00026
#2	.00309	.78404	.00069	.00619	.01142	.00023	9.4483	-.00028
#3	.00480	.77489	-.00084	.00682	.00960	.00016	9.2678	-.00014
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00067**	**.00209**	**.00225**	**1.1046**	**.20587**	**.00183**	**3.8329**	**.12468**
Stddev	.00025	.00124	.00151	.0075	.05413	.00323	.0374	.00185
%RSD	36.537	59.143	66.907	.67811	26.294	176.78	.97568	1.4875
#1	.00057	.00084	.00284	1.1048	.18555	.00515	3.8178	.12286
#2	.00096	.00331	.00337	1.1121	.26722	.00166	3.8755	.12657
#3	.00050	.00212	.00054	1.0971	.16483	-.00131	3.8054	.12461
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00011**	**.91553**	**.00104**	**.04284**	**.00087**	**-.00434**	**.00416**	**1.4218**
Stddev	.00049	.02558	.00097	.00374	.00128	.00145	.00383	.0304
%RSD	448.71	2.7944	92.896	8.7405	147.22	33.355	92.023	2.1368
#1	.00043	.88763	.00204	.04714	.00040	-.00289	.00799	1.3901
#2	-.00054	.92105	.00011	.04033	-.00011	-.00435	.00416	1.4246
#3	-.00022	.93790	.00097	.04105	.00232	-.00579	.00033	1.4507
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127902 Acquired: 6/25/2016 15:25:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00026**	**.04596**	**.00940**	**.00124**	**.00257**	**.00907**	**.00122**
Stddev	.00038	.00038	.00559	.00152	.00046	.00027	.00089
%RSD	147.88	.81748	59.487	122.54	17.946	3.0009	73.195
#1	.00017	.04588	.00305	-.00019	.00219	.00893	.00224
#2	-.00038	.04637	.01155	.00284	.00243	.00890	.00056
#3	-.00056	.04564	.01360	.00108	.00308	.00938	.00087
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10039.**	**61079.**	**3063.3**
Stddev	16.	87.	16.8
%RSD	.15707	.14311	.54820
#1	10022.	60979.	3046.2
#2	10041.	61122.	3064.0
#3	10054.	61137.	3079.8

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 15:29:11 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.41474	10.480	.41122	.51129	1.0445	.05223	10.384
Stddev	.00157	.008	.00187	.00319	.0019	.00016	.024
%RSD	.37799	.07401	.45408	.62455	.18188	.31144	.22798
#1	.41325	10.484	.41182	.51035	1.0465	.05219	10.387
#2	.41461	10.484	.41272	.50866	1.0441	.05241	10.359
#3	.41637	10.471	.40913	.51484	1.0428	.05209	10.407
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05244	.20643	.51777	.52807	4.1528	50.626	1.0452
Stddev	.00019	.00026	.00163	.00025	.0357	.257	.0080
%RSD	.36313	.12673	.31564	.04665	.85968	.50669	.76446
#1	.05244	.20669	.51592	.52825	4.1190	50.704	1.0431
#2	.05226	.20644	.51838	.52817	4.1901	50.835	1.0540
#3	.05264	.20617	.51902	.52779	4.1492	50.340	1.0385
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.9629	.52245	1.0531	51.247	.52506	10.312	.52216
Stddev	.0274	.00433	.0047	.134	.00127	.013	.00461
%RSD	.27495	.82847	.44799	.26095	.24264	.12919	.88203
#1	9.9944	.51946	1.0578	51.179	.52653	10.320	.51773
#2	9.9492	.52742	1.0532	51.401	.52432	10.296	.52692
#3	9.9451	.52049	1.0484	51.161	.52433	10.319	.52182
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 15:29:11 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2356**	**.41239**	**5.3381**	**1.0346**	**1.0460**	**1.0497**	**.53160**
Stddev	.0017	.00642	.0057	.0013	.0038	.0052	.00142
%RSD	.13528	1.5556	.10675	.12483	.36598	.49394	.26637
#1	1.2362	.41861	5.3408	1.0357	1.0443	1.0445	.53322
#2	1.2337	.40580	5.3420	1.0349	1.0504	1.0497	.53096
#3	1.2369	.41277	5.3316	1.0332	1.0433	1.0549	.53062
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0558**	**1.0420**	F **3.0350**
Stddev	.0013	.0007	.0090
%RSD	.12612	.06683	.29605
#1	1.0548	1.0423	3.0277
#2	1.0553	1.0412	3.0451
#3	1.0573	1.0426	3.0323
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9586.8**	**58077.**	**2975.5**
Stddev	6.0	176.	6.3
%RSD	.06238	.30386	.21060
#1	9587.3	57874.	2969.1
#2	9592.5	58193.	2975.6
#3	9580.6	58165.	2981.7

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 15:32:48 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00000**	**.00002**	**-.00010**	**.00491**	**.00210**	**.00011**	**.00565**	**-.00003**
Stddev	.00087	.00719	.00078	.00229	.00030	.00005	.00049	.00029
%RSD	26249.	33697.	808.92	46.626	14.233	46.346	8.5911	909.94
#1	.00025	.00801	-.00086	.00753	.00176	.00017	.00542	-.00000
#2	.00073	-.00593	.00069	.00390	.00223	.00007	.00533	-.00034
#3	-.00097	-.00201	-.00012	.00330	.00231	.00010	.00621	.00024
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00020**	**.00011**	**.00064**	**.03500**	**.03424**	**-.00179**	**.02594**	**-.00052**
Stddev	.00009	.00046	.00051	.01471	.03973	.00240	.07741	.00156
%RSD	43.655	413.87	80.028	42.019	116.03	134.20	298.41	302.41
#1	.00012	-.00018	.00036	.01804	-.01111	-.00218	.05944	-.00155
#2	.00020	.00064	.00123	.04417	.05092	-.00397	.08095	.00128
#3	.00029	-.00013	.00033	.04280	.06290	.00078	-.06257	-.00128
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00316**	**.04507**	**.00037**	**.00106**	**.00080**	**.00374**	**.00412**	**.00990**
Stddev	.00041	.02068	.00057	.00881	.00149	.00197	.00171	.00085
%RSD	13.022	45.885	154.95	828.64	185.93	52.502	41.416	8.5529
#1	.00281	.06772	-.00005	.00994	-.00040	.00294	.00217	.01067
#2	.00361	.02718	.00102	.00092	.00033	.00230	.00488	.01002
#3	.00306	.04031	.00014	-.00767	.00246	.00598	.00532	.00899
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 15:32:48 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00048**	**.00057**	**.00333**	**-.00225**	**.00169**	**-.00006**	**.00439**
Stddev	.00073	.00051	.00753	.00264	.00040	.00020	.00135
%RSD	152.10	89.662	226.48	117.22	23.766	306.69	30.851
#1	.00129	.00079	-.00329	-.00367	.00165	-.00017	.00283
#2	-.00014	-.00001	.01153	-.00387	.00130	-.00018	.00508
#3	.00029	.00092	.00174	.00079	.00210	.00016	.00526
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9759.3**	**59359.**	**3013.9**
Stddev	2.3	121.	1.6
%RSD	.02366	.20394	.05186
#1	9761.6	59292.	3015.4
#2	9759.3	59287.	3013.9
#3	9756.9	59499.	3012.3

Approved: June 27, 2016

Sample Name: L1606127903S Acquired: 6/25/2016 15:36:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01197	1.1036	.01087	.05712	.03304	.00134	9.9729	.00124
Stddev	.00119	.0059	.00187	.00108	.00116	.00008	.0996	.00005
%RSD	9.9297	.53710	17.218	1.8903	3.5218	6.1377	.99838	3.7414
#1	.01115	1.1017	.01246	.05619	.03414	.00126	10.019	.00125
#2	.01142	1.1103	.00881	.05688	.03182	.00135	9.8587	.00119
#3	.01333	1.0989	.01134	.05830	.03317	.00142	10.041	.00127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00560	.01397	.01462	1.0767	1.4208	.02780	4.2261	.14457
Stddev	.00014	.00108	.00214	.0241	.1192	.00407	.0440	.00290
%RSD	2.4102	7.7598	14.602	2.2369	8.3867	14.648	1.0401	2.0070
#1	.00574	.01498	.01254	1.0620	1.3125	.02932	4.2372	.14280
#2	.00559	.01283	.01680	1.0635	1.4014	.02319	4.1777	.14299
#3	.00547	.01411	.01453	1.1045	1.5484	.03090	4.2635	.14792
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02152	2.1329	.01322	.29583	.01584	.02588	.00942	1.7458
Stddev	.00098	.0418	.00158	.00727	.00355	.00244	.00136	.0125
%RSD	4.5315	1.9579	11.965	2.4574	22.390	9.4200	14.478	.71438
#1	.02231	2.1343	.01139	.30417	.01767	.02397	.01094	1.7523
#2	.02182	2.0905	.01413	.29245	.01809	.02862	.00829	1.7537
#3	.02043	2.1740	.01413	.29086	.01175	.02503	.00905	1.7314
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127903S Acquired: 6/25/2016 15:36:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02137**	**.07160**	**.03468**	**.01383**	**.02654**	**.02985**	**.00242**
Stddev	.00036	.00081	.00438	.00291	.00046	.00010	.00158
%RSD	1.6937	1.1317	12.615	21.070	1.7354	.32151	65.388
#1	.02178	.07099	.02963	.01549	.02613	.02986	.00141
#2	.02110	.07130	.03735	.01047	.02704	.02975	.00424
#3	.02123	.07252	.03706	.01554	.02644	.02994	.00160
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9975.1**	**60274.**	**3041.7**
Stddev	49.0	118.	22.7
%RSD	.49139	.19506	.74582
#1	9940.1	60350.	3061.9
#2	9954.2	60332.	3046.1
#3	10031.	60138.	3017.2

Approved: June 27, 2016

Sample Name: L1606127904SD Acquired: 6/25/2016 15:40:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.01103**	**1.1028**	**.01203**	**.05464**	**.03364**	**.00143**	**9.8642**	**.00121**
Stddev	.00186	.0090	.00180	.00231	.00086	.00004	.1305	.00006
%RSD	16.883	.81742	14.963	4.2286	2.5577	2.6326	1.3231	5.3301
#1	.00912	1.0960	.01267	.05217	.03303	.00146	9.7151	.00122
#2	.01113	1.1130	.00999	.05675	.03463	.00143	9.9198	.00114
#3	.01284	1.0994	.01342	.05499	.03327	.00139	9.9577	.00127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00590**	**.01420**	**.01484**	**1.0576**	**1.4598**	**.02493**	**4.2188**	**.14616**
Stddev	.00020	.00048	.00068	.0082	.1286	.00072	.1177	.00241
%RSD	3.3297	3.4140	4.5711	.77893	8.8104	2.9051	2.7895	1.6455
#1	.00602	.01364	.01562	1.0663	1.6059	.02466	4.0843	.14411
#2	.00567	.01444	.01454	1.0499	1.4099	.02439	4.2697	.14557
#3	.00601	.01451	.01437	1.0566	1.3636	.02576	4.3025	.14881
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02217**	**2.1040**	**.01418**	**.30006**	**.01556**	**.02408**	**.01453**	**1.7577**
Stddev	.00072	.0376	.00057	.00408	.00448	.00503	.00712	.0151
%RSD	3.2386	1.7891	4.0074	1.3597	28.796	20.896	49.023	.85907
#1	.02294	2.0606	.01353	.29698	.01280	.02854	.02252	1.7414
#2	.02203	2.1275	.01453	.30469	.02073	.02509	.00885	1.7604
#3	.02153	2.1239	.01449	.29851	.01315	.01862	.01221	1.7712
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127904SD Acquired: 6/25/2016 15:40:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573994-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02276**	**.07116**	**.03577**	**.01268**	**.02759**	**.03003**	**.00230**
Stddev	.00014	.00090	.00737	.00234	.00032	.00029	.00174
%RSD	.61715	1.2587	20.620	18.483	1.1623	.96154	75.576
#1	.02272	.07054	.03703	.01291	.02794	.02969	.00432
#2	.02292	.07076	.02784	.01490	.02752	.03021	.00132
#3	.02264	.07219	.04242	.01023	.02731	.03018	.00128
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10006.**	**60810.**	**3063.3**
Stddev	5.	192.	18.4
%RSD	.04992	.31626	.60150
#1	10009.	61010.	3082.6
#2	10009.	60627.	3061.4
#3	10000.	60793.	3045.9

Approved: June 27, 2016

Sample Name: L1606128901 Acquired: 6/25/2016 15:44:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00215	.00687	.00324	.00036	.25940	.00003	8.5369	.00005
Stddev	.00051	.01020	.00141	.00225	.00475	.00006	.1296	.00011
%RSD	23.677	148.57	43.630	628.60	1.8323	170.08	1.5176	227.30
#1	.00216	.01614	.00358	-.00039	.25471	.00001	8.3890	.00017
#2	.00265	.00852	.00169	.00289	.26421	.00010	8.6303	-.00003
#3	.00163	-.00406	.00445	-.00143	.25928	-.00000	8.5913	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00259	.00588	.00175	.08631	.04218	.00151	6.5072	.07542
Stddev	.00040	.00134	.00073	.02244	.04887	.00193	.2142	.00074
%RSD	15.316	22.711	41.571	26.002	115.87	127.81	3.2912	.97595
#1	.00214	.00436	.00096	.06469	.09555	-.00070	6.2873	.07459
#2	.00279	.00683	.00239	.10949	-.00039	.00286	6.5191	.07570
#3	.00285	.00647	.00190	.08474	.03137	.00237	6.7151	.07598
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	23.216	.03092	.00614	.00300	-.00111	.00640	.97050
Stddev	.00039	.461	.00078	.00363	.00087	.00404	.00265	.00247
%RSD	280.63	1.9852	2.5231	59.153	29.021	362.80	41.403	.25451
#1	-.00028	22.689	.03110	.00526	.00208	-.00358	.00616	.97332
#2	.00048	23.539	.03160	.01013	.00311	-.00330	.00916	.96949
#3	.00022	23.422	.03007	.00303	.00381	.00355	.00387	.96870
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606128901 Acquired: 6/25/2016 15:44:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00051**	**.16522**	**-.00360**	**-.00090**	**.00125**	**.00115**	**.00088**
Stddev	.00057	.00251	.00574	.00300	.00085	.00009	.00100
%RSD	111.15	1.5196	159.60	334.58	68.243	8.1341	114.01
#1	-.00005	.16233	-.01006	-.00433	.00031	.00125	.00060
#2	.00108	.16651	.00089	.00117	.00197	.00107	.00199
#3	.00050	.16682	-.00162	.00047	.00148	.00112	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9940.0**	**60088.**	**3038.6**
Stddev	4.2	150.	44.1
%RSD	.04218	.24912	1.4500
#1	9935.3	60256.	3084.8
#2	9941.7	60040.	2997.0
#3	9943.1	59969.	3034.2

Approved: June 27, 2016

Sample Name: L1606112901 Acquired: 6/25/2016 15:48:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00126	-.00046	-.00005	-.00065	.00318	.00006	4.1415	-.00016
Stddev	.00197	.00573	.00201	.00160	.00151	.00005	.0124	.00023
%RSD	156.91	1238.2	4274.9	245.00	47.612	81.871	.30015	142.20
#1	.00341	-.00015	-.00215	-.00141	.00362	.00006	4.1534	-.00032
#2	-.00045	.00511	.00185	.00118	.00150	.00012	4.1425	.00010
#3	.00081	-.00635	.00016	-.00173	.00443	.00001	4.1286	-.00026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00001	.00058	.00066	.05376	.10451	-.00223	.92501	.00486
Stddev	.00022	.00057	.00151	.01103	.06065	.00348	.08081	.00144
%RSD	3258.5	98.481	227.65	20.514	58.030	156.17	8.7365	29.653
#1	-.00025	.00121	.00094	.05934	.12513	.00172	.91232	.00510
#2	.00010	.00039	-.00097	.04105	.03624	-.00357	1.0114	.00331
#3	.00016	.00012	.00201	.06088	.15216	-.00484	.85130	.00616
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00027	3.0376	-.00008	.00578	.00044	-.00230	.00144	.21434
Stddev	.00025	.0744	.00053	.00529	.00227	.00178	.00420	.00577
%RSD	93.741	2.4506	639.89	91.432	509.94	77.275	291.44	2.6916
#1	-.00011	2.9794	.00051	.00238	.00289	-.00166	.00573	.20865
#2	-.00014	3.0119	-.00050	.00310	-.00159	-.00431	-.00267	.21418
#3	-.00056	3.1215	-.00026	.01187	.00003	-.00093	.00126	.22019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606112901 Acquired: 6/25/2016 15:48:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00020**	**.01265**	**.00270**	**.00064**	**.00099**	**.00067**	**.00092**
Stddev	.00055	.00035	.00371	.00238	.00052	.00018	.00051
%RSD	272.21	2.7272	137.50	373.62	52.293	26.328	55.163
#1	-.00083	.01259	.00376	-.00183	.00159	.00059	.00050
#2	.00008	.01234	-.00143	.00082	.00070	.00087	.00077
#3	.00015	.01303	.00576	.00292	.00069	.00054	.00148
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9969.3**	**60775.**	**3044.6**
Stddev	21.9	129.	15.2
%RSD	.22015	.21273	.49841
#1	9984.3	60920.	3056.8
#2	9944.1	60732.	3049.4
#3	9979.5	60672.	3027.6

Approved: June 27, 2016

Sample Name: L1606112902 Acquired: 6/25/2016 15:52:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00026	.00804	.00035	-.00128	.00288	.00010	2.9594	-.00006
Stddev	.00058	.00641	.00226	.00137	.00187	.00003	.0055	.00010
%RSD	221.07	79.747	653.27	107.46	64.882	33.893	.18419	164.92
#1	.00004	.00818	.00185	.00031	.00398	.00007	2.9654	.00000
#2	.00011	.00156	.00144	-.00200	.00072	.00009	2.9582	-.00001
#3	-.00094	.01439	-.00226	-.00213	.00394	.00014	2.9547	-.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00001	.00045	.00076	.08667	.10793	.00221	.95476	.02200
Stddev	.00021	.00074	.00096	.00189	.05164	.00058	.08086	.00168
%RSD	3252.3	165.35	125.16	2.1842	47.845	26.260	8.4687	7.6363
#1	.00009	.00055	.00166	.08573	.06794	.00282	.86999	.02017
#2	-.00024	-.00034	.00088	.08885	.16623	.00215	.96327	.02346
#3	.00016	.00114	-.00024	.08543	.08961	.00166	1.0310	.02238
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00025	1.4239	.00014	.00542	.00025	-.00059	.00178	.20161
Stddev	.00013	.0489	.00086	.00331	.00120	.00643	.00247	.00182
%RSD	52.086	3.4352	596.95	60.989	470.84	1086.5	138.88	.90303
#1	-.00010	1.4077	-.00013	.00438	-.00113	.00173	.00080	.20089
#2	-.00031	1.4789	-.00055	.00913	.00091	-.00786	-.00005	.20368
#3	-.00033	1.3852	.00110	.00276	.00098	.00435	.00459	.20026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606112902 Acquired: 6/25/2016 15:52:22 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00083	.01066	.00349	-.00281	.00049	.00074	-.00046
Stddev	.00025	.00027	.00348	.00166	.00046	.00011	.00023
%RSD	30.685	2.4916	99.744	59.148	94.747	14.382	50.877
#1	.00087	.01070	.00643	-.00461	.00075	.00069	-.00046
#2	.00056	.01090	.00439	-.00132	-.00005	.00087	-.00023
#3	.00106	.01038	-.00035	-.00252	.00076	.00067	-.00070
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9989.3	60825.	3066.0
Stddev	18.4	89.	4.0
%RSD	.18402	.14693	.13203
#1	10008.	60870.	3065.9
#2	9971.8	60722.	3062.0
#3	9987.6	60883.	3070.1

Approved: June 27, 2016

Sample Name: L1606112903 Acquired: 6/25/2016 15:56:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00034**	**-.00746**	**.00009**	**.00007**	**.00214**	**.00016**	**4.2282**	**-.00031**
Stddev	.00178	.00543	.00185	.00098	.00088	.00001	.1021	.00011
%RSD	518.14	72.757	1962.2	1427.6	41.098	3.6332	2.4147	36.326
#1	-.00212	-.00132	-.00121	.00003	.00116	.00017	4.2310	-.00019
#2	.00143	-.00942	.00221	-.00089	.00287	.00016	4.3288	-.00035
#3	-.00034	-.01164	-.00072	.00107	.00238	.00015	4.1247	-.00041
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00027**	**-.00057**	**-.00029**	**.07902**	**-.03881**	**.00121**	**1.0358**	**.00397**
Stddev	.00063	.00094	.00019	.01076	.01948	.00237	.1206	.00030
%RSD	236.86	166.12	66.025	13.622	50.186	195.96	11.648	7.4974
#1	-.00020	-.00113	-.00052	.06755	-.06059	-.00026	.93657	.00379
#2	.00099	.00052	-.00021	.08890	-.02305	-.00005	1.1701	.00431
#3	.00002	-.00109	-.00016	.08063	-.03281	.00395	1.0007	.00380
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00055**	**2.1503**	**-.00065**	**.00144**	**.00315**	**-.00171**	**-.00059**	**.27406**
Stddev	.00026	.0349	.00083	.00721	.00266	.00193	.00290	.00085
%RSD	47.062	1.6229	127.58	500.84	84.421	112.70	495.02	.31027
#1	-.00082	2.1700	-.00132	-.00288	.00271	.00020	-.00182	.27322
#2	-.00053	2.1710	.00028	-.00257	.00074	-.00366	.00273	.27492
#3	-.00030	2.1101	-.00091	.00976	.00599	-.00168	-.00267	.27403
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606112903 Acquired: 6/25/2016 15:56:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	.01267	.00761	-.00035	.00044	.00047	.00034
Stddev	.00036	.00010	.00488	.00192	.00076	.00017	.00099
%RSD	75.051	.78508	64.089	556.65	172.53	36.380	290.00
#1	.00045	.01256	.01172	-.00046	-.00009	.00039	-.00046
#2	.00086	.01267	.00891	.00163	.00010	.00066	.00145
#3	.00013	.01276	.00222	-.00221	.00131	.00034	.00003
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9986.0	60665.	3026.9
Stddev	20.7	145.	17.8
%RSD	.20708	.23844	.58822
#1	10007.	60770.	3014.6
#2	9984.8	60726.	3018.9
#3	9966.0	60500.	3047.3

Approved: June 27, 2016

Sample Name: L1606112904 Acquired: 6/25/2016 16:00:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00143	.00782	.00014	-.00105	.00318	.00018	2.9065	-.00009
Stddev	.00049	.00478	.00214	.00233	.00102	.00003	.0438	.00019
%RSD	34.335	61.153	1520.7	221.14	32.071	15.659	1.5085	207.21
#1	.00199	.01026	-.00194	.00103	.00202	.00019	2.8571	-.00015
#2	.00122	.00231	.00002	-.00063	.00358	.00019	2.9407	.00012
#3	.00108	.01088	.00233	-.00357	.00394	.00014	2.9218	-.00024
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00050	.00045	.00112	.05226	.04902	.00077	.91311	.02436
Stddev	.00068	.00160	.00112	.00414	.07439	.00531	.01971	.00311
%RSD	135.32	357.44	99.972	7.9272	151.77	691.64	2.1590	12.769
#1	.00007	.00219	.00019	.05089	.09310	-.00511	.93142	.02095
#2	.00129	.00012	.00080	.04898	.09083	.00221	.91566	.02510
#3	.00015	-.00096	.00236	.05692	-.03687	.00521	.89224	.02704
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00023	1.3676	-.00081	.00778	.00502	-.00289	.00478	.20680
Stddev	.00041	.0202	.00063	.00381	.00153	.00221	.00206	.00531
%RSD	176.12	1.4731	77.366	49.002	30.478	76.529	43.075	2.5661
#1	-.00056	1.3492	-.00140	.01136	.00503	-.00266	.00633	.20082
#2	-.00037	1.3644	-.00015	.00819	.00655	-.00520	.00556	.20859
#3	.00023	1.3891	-.00088	.00377	.00348	-.00080	.00244	.21097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606112904 Acquired: 6/25/2016 16:00:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00019	.01047	.00111	-.00019	.00078	.00080	-.00006
Stddev	.00022	.00018	.00665	.00270	.00041	.00022	.00073
%RSD	116.67	1.7478	597.28	1390.3	52.913	27.125	1123.9
#1	.00044	.01026	.00785	.00028	.00121	.00056	.00012
#2	.00008	.01056	.00093	.00224	.00075	.00085	.00055
#3	.00005	.01059	-.00544	-.00310	.00038	.00099	-.00086
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9957.7	60557.	3030.1
Stddev	15.8	177.	8.6
%RSD	.15901	.29213	.28426
#1	9975.9	60586.	3020.4
#2	9949.3	60718.	3033.0
#3	9947.7	60368.	3036.9

Approved: June 27, 2016

Sample Name: L1606127201 Acquired: 6/25/2016 16:04:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00028	-.00887	-.00183	.00197	.00056	.00010	.95406	-.00003
Stddev	.00091	.01117	.00052	.00188	.00043	.00011	.05017	.00010
%RSD	325.46	125.87	28.528	95.488	77.565	107.58	5.2586	289.02
#1	.00022	-.02176	-.00218	.00027	.00085	.00010	.90480	.00002
#2	.00122	-.00213	-.00123	.00398	.00076	-.00001	.95228	.00003
#3	-.00060	-.00272	-.00207	.00165	.00006	.00021	1.0051	-.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00003	-.00128	.00046	.02799	-.07233	.00197	.21481	.00019
Stddev	.00061	.00045	.00022	.01180	.06961	.00022	.12156	.00200
%RSD	1838.2	35.569	48.564	42.171	96.239	11.394	56.590	1042.3
#1	-.00013	-.00080	.00028	.03989	-.13804	.00212	.34118	.00130
#2	-.00048	-.00170	.00071	.01628	.00061	.00171	.09871	-.00212
#3	.00071	-.00133	.00040	.02780	-.07955	.00208	.20454	.00139
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00067	6.5927	-.00041	.00599	.00226	-.00187	.00139	.23314
Stddev	.00039	.0563	.00074	.00538	.00192	.00575	.00169	.00199
%RSD	58.032	.85407	180.55	89.733	84.902	308.02	122.23	.85400
#1	-.00069	6.5318	-.00010	.00974	.00276	.00008	.00216	.23308
#2	-.00105	6.6428	-.00125	-.00017	.00014	-.00834	-.00056	.23517
#3	-.00027	6.6034	.00012	.00842	.00389	.00266	.00255	.23118
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127201 Acquired: 6/25/2016 16:04:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00023**	**.00331**	**.00248**	**.00096**	**.00107**	**.00085**	**.00133**
Stddev	.00044	.00056	.00254	.00297	.00078	.00012	.00063
%RSD	192.52	16.945	102.26	309.29	73.373	14.693	47.329
#1	.00017	.00395	.00360	.00246	.00196	.00094	.00062
#2	-.00070	.00291	.00428	.00288	.00074	.00089	.00180
#3	-.00016	.00306	-.00042	-.00246	.00050	.00071	.00159
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9972.6**	**60217.**	**3037.7**
Stddev	14.0	83.	14.1
%RSD	.14049	.13799	.46569
#1	9988.5	60309.	3026.4
#2	9967.3	60146.	3033.1
#3	9962.1	60196.	3053.5

Approved: June 27, 2016

Sample Name: L1606127202 Acquired: 6/25/2016 16:08:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00026**	**.01060**	**.00084**	**.01591**	**.00482**	**.00013**	**4.8874**	**.00012**
Stddev	.00147	.00116	.00295	.00196	.00045	.00002	.0512	.00031
%RSD	570.05	10.898	350.53	12.303	9.3008	12.581	1.0477	272.03
#1	-.00187	.01061	.00261	.01812	.00447	.00014	4.8299	.00013
#2	.00100	.00945	-.00256	.01523	.00533	.00015	4.9279	.00042
#3	.00009	.01176	.00248	.01439	.00467	.00011	4.9046	-.00021
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00019**	**-.00046**	**.00025**	**.03059**	**.03395**	**.00208**	**2.7028**	**-.00061**
Stddev	.00015	.00087	.00083	.01638	.06594	.00230	.1335	.00149
%RSD	81.112	189.49	337.36	53.558	194.25	110.69	4.9393	243.32
#1	.00003	-.00146	.00109	.04846	-.00968	-.00038	2.8372	-.00173
#2	.00034	.00002	-.00057	.01628	.00172	.00418	2.5702	.00108
#3	.00021	.00006	.00022	.02703	.10981	.00243	2.7010	-.00120
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00026**	**13.742**	**-.00085**	**.00227**	**.00279**	**-.00269**	**-.00193**	**.47645**
Stddev	.00056	.155	.00030	.00745	.00217	.00177	.00170	.00665
%RSD	217.74	1.1297	35.156	328.49	77.588	65.958	88.159	1.3954
#1	-.00002	13.587	-.00065	-.00435	.00526	-.00088	-.00382	.46927
#2	-.00089	13.898	-.00120	.00082	.00118	-.00276	-.00050	.47768
#3	.00014	13.742	-.00071	.01033	.00194	-.00443	-.00148	.48240
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127202 Acquired: 6/25/2016 16:08:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00018**	**.05092**	**.00177**	**.00024**	**.00137**	**.00071**	**.00162**
Stddev	.00083	.00100	.00998	.00193	.00052	.00009	.00084
%RSD	474.40	1.9658	565.04	789.52	37.663	12.612	51.686
#1	-.00049	.05003	.00781	-.00019	.00196	.00074	.00252
#2	-.00081	.05200	-.00975	.00235	.00120	.00078	.00148
#3	.00077	.05072	.00724	-.00143	.00097	.00061	.00086
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9920.3**	**59957.**	**3053.1**
Stddev	9.3	97.	28.1
%RSD	.09367	.16197	.92019
#1	9910.9	59874.	3056.5
#2	9929.5	60064.	3023.5
#3	9920.6	59932.	3079.4

Approved: June 27, 2016

Sample Name: L1606127901 Acquired: 6/25/2016 16:12:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00079	.23343	-.00035	.00697	.01048	.00010	5.5956	-.00004
Stddev	.00151	.00808	.00093	.00046	.00038	.00003	.0897	.00007
%RSD	191.07	3.4604	268.33	6.6127	3.6255	27.106	1.6038	155.88
#1	-.00012	.24183	.00073	.00738	.01041	.00014	5.5127	-.00007
#2	.00027	.23274	-.00087	.00647	.01090	.00009	5.5834	-.00010
#3	-.00251	.22572	-.00090	.00704	.01015	.00009	5.6909	.00003
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00030	.00040	.00104	.37949	.02999	.00120	2.2214	.04375
Stddev	.00022	.00063	.00117	.02000	.03448	.00297	.1047	.00042
%RSD	72.888	155.32	112.12	5.2704	114.98	248.44	4.7142	.95613
#1	.00020	.00102	-.00030	.35646	.01685	.00172	2.1668	.04335
#2	.00015	-.00023	.00181	.39248	.06911	.00387	2.1552	.04372
#3	.00056	.00042	.00162	.38953	.00401	-.00200	2.3421	.04419
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00013	.40372	.00002	.01800	.00182	-.00229	.00534	.76285
Stddev	.00016	.02832	.00029	.00724	.00294	.00133	.00501	.02556
%RSD	124.97	7.0152	1295.9	40.230	161.43	58.036	93.812	3.3508
#1	-.00031	.38955	.00035	.01067	.00373	-.00090	.01054	.73582
#2	-.00001	.38528	-.00008	.01820	-.00157	-.00355	.00492	.76609
#3	-.00007	.43633	-.00020	.02515	.00330	-.00243	.00055	.78664
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127901 Acquired: 6/25/2016 16:12:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00000**	**.02802**	**.00639**	**-.00424**	**.00104**	**.00686**	**-.00053**
Stddev	.00059	.00084	.00088	.00033	.00039	.00027	.00162
%RSD	12664.	3.0023	13.821	7.8780	37.381	4.0081	307.76
#1	.00044	.02706	.00703	-.00390	.00138	.00661	.00100
#2	.00022	.02835	.00538	-.00425	.00062	.00681	-.00222
#3	-.00068	.02864	.00675	-.00456	.00114	.00715	-.00036
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9995.1**	**60569.**	**3072.5**
Stddev	18.8	83.	17.9
%RSD	.18798	.13702	.58354
#1	10017.	60498.	3093.1
#2	9983.0	60660.	3060.2
#3	9985.6	60548.	3064.2

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 16:16:03 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41279**	**10.433**	**.40298**	**.50323**	**1.0184**	**.05236**	**10.146**
Stddev	.00062	.023	.00491	.00242	.0037	.00006	.030
%RSD	.14995	.22399	1.2186	.48043	.36729	.12018	.29415
#1	.41231	10.415	.40384	.50055	1.0220	.05229	10.131
#2	.41349	10.459	.39769	.50390	1.0146	.05241	10.126
#3	.41257	10.425	.40740	.50524	1.0186	.05238	10.180
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05206**	**.20294**	**.51645**	**.52099**	**4.0952**	**49.051**	**1.0295**
Stddev	.00043	.00065	.00457	.00121	.0025	.142	.0088
%RSD	.83054	.31937	.88415	.23159	.06120	.28862	.85476
#1	.05248	.20360	.51186	.52157	4.0943	49.074	1.0305
#2	.05210	.20291	.52099	.52180	4.0933	48.900	1.0203
#3	.05161	.20230	.51649	.51960	4.0980	49.180	1.0378
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.7277**	**.51414**	**1.0402**	**49.935**	**.51586**	**10.134**	**.51189**
Stddev	.0462	.00321	.0047	.095	.00165	.031	.00166
%RSD	.47447	.62359	.45388	.19063	.31918	.30495	.32486
#1	9.7465	.51782	1.0451	50.021	.51766	10.168	.51381
#2	9.6751	.51192	1.0398	49.833	.51549	10.123	.51086
#3	9.7615	.51270	1.0357	49.951	.51442	10.109	.51101
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 16:16:03 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2153**	**.40661**	**5.2188**	**1.0211**	**1.0187**	**1.0271**	**.52415**
Stddev	.0058	.00386	.0120	.0013	.0031	.0069	.00046
%RSD	.47598	.95000	.22957	.12690	.30817	.67428	.08834
#1	1.2162	.40345	5.2255	1.0218	1.0207	1.0286	.52363
#2	1.2206	.41091	5.2258	1.0219	1.0150	1.0195	.52433
#3	1.2091	.40546	5.2049	1.0196	1.0203	1.0331	.52450
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0611**	**1.0269**	F **2.9985**
Stddev	.0021	.0010	.0124
%RSD	.19830	.09638	.41405
#1	1.0631	1.0270	3.0047
#2	1.0614	1.0279	2.9842
#3	1.0589	1.0259	3.0065
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9559.6**	**57599.**	**2986.8**
Stddev	4.9	130.	10.5
%RSD	.05148	.22635	.35191
#1	9564.0	57612.	2977.9
#2	9560.6	57462.	2998.4
#3	9554.3	57722.	2984.0

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 16:19:40 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00185	.00626	-.00023	.00044	.00278	.00005	-.04464	.00001
Stddev	.00025	.00998	.00075	.00197	.00045	.00006	.01589	.00008
%RSD	13.751	159.28	328.05	451.88	16.156	122.09	35.589	834.92
#1	.00158	.01229	-.00019	.00230	.00236	.00010	-.04122	-.00004
#2	.00190	-.00525	.00050	.00063	.00326	-.00002	-.03074	-.00004
#3	.00208	.01176	-.00099	-.00162	.00274	.00006	-.06195	.00010
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00039	.00030	.00108	.01891	-.08883	.00028	.04170	-.00007
Stddev	.00014	.00149	.00123	.02463	.04133	.00158	.04172	.00069
%RSD	37.231	490.44	114.38	130.22	46.533	569.52	100.03	937.33
#1	-.00023	.00135	.00225	.04102	-.10527	-.00061	.02819	.00053
#2	-.00043	.00096	-.00021	.02335	-.11941	.00210	.00842	.00008
#3	-.00050	-.00140	.00119	-.00763	-.04180	-.00066	.08850	-.00082
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00343	.00251	-.00012	.00363	.00401	.00489	.00879	-.00487
Stddev	.00038	.01645	.00050	.00608	.00167	.00193	.00298	.00028
%RSD	11.063	654.49	431.61	167.28	41.713	39.486	33.953	5.7618
#1	.00299	.00408	-.00004	.00277	.00588	.00378	.00970	-.00459
#2	.00370	-.01467	-.00065	-.00196	.00352	.00712	.01122	-.00515
#3	.00359	.01813	.00034	.01010	.00264	.00378	.00546	-.00487
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 16:19:40 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00028**	**.00057**	**.00214**	**-.00157**	**.00060**	**-.00006**	**.00421**
Stddev	.00055	.00026	.00424	.00416	.00034	.00004	.00038
%RSD	193.42	45.468	198.11	264.68	55.851	78.634	9.1291
#1	-.00056	.00032	.00612	.00235	.00061	-.00008	.00456
#2	-.00063	.00083	-.00232	-.00593	.00093	-.00001	.00426
#3	.00035	.00056	.00262	-.00114	.00026	-.00008	.00380
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9751.1**	**59011.**	**2998.0**
Stddev	16.0	92.	16.9
%RSD	.16409	.15669	.56327
#1	9735.8	59057.	2997.7
#2	9767.8	59072.	2981.2
#3	9749.7	58905.	3015.0

Approved: June 27, 2016

Sample Name: L1606127905 Acquired: 6/25/2016 16:23:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00112	.42892	.00254	.00885	.00931	.00014	4.1101	-.00014
Stddev	.00085	.00960	.00110	.00026	.00054	.00008	.0226	.00009
%RSD	76.091	2.2373	43.205	2.9736	5.7541	61.416	.54879	60.631
#1	.00187	.41894	.00336	.00888	.00991	.00015	4.0848	-.00017
#2	.00019	.43809	.00129	.00858	.00890	.00022	4.1178	-.00005
#3	.00131	.42973	.00298	.00910	.00911	.00005	4.1279	-.00021
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00064	.00072	.00093	.46806	-.04215	.00375	2.3175	.01534
Stddev	.00025	.00122	.00082	.01630	.03381	.00237	.0180	.00171
%RSD	38.393	169.05	88.133	3.4825	80.210	63.236	.77530	11.156
#1	.00092	.00051	.00011	.45372	-.08111	.00533	2.3322	.01684
#2	.00057	-.00038	.00092	.46467	-.02046	.00102	2.2974	.01572
#3	.00044	.00203	.00175	.48579	-.02489	.00491	2.3228	.01348
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00042	.79830	.00036	.03423	.00164	-.00327	.00568	1.0453
Stddev	.00055	.00874	.00054	.00334	.00066	.00436	.00399	.0205
%RSD	130.46	1.0949	148.73	9.7489	39.962	133.66	70.190	1.9601
#1	-.00012	.80324	-.00001	.03061	.00234	-.00815	.00802	1.0252
#2	.00040	.78821	.00099	.03719	.00155	-.00191	.00794	1.0445
#3	.00099	.80344	.00012	.03489	.00104	.00026	.00108	1.0662
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606127905 Acquired: 6/25/2016 16:23:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.01799	.00482	-.00230	.00189	.01109	.00007
Stddev	.00065	.00037	.00226	.00232	.00074	.00029	.00211
%RSD	715.49	2.0334	47.005	100.92	39.126	2.6113	3180.6
#1	-.00003	.01830	.00310	.00003	.00148	.01082	.00082
#2	.00080	.01809	.00396	-.00232	.00274	.01107	-.00231
#3	-.00049	.01759	.00738	-.00462	.00144	.01140	.00169
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9972.5	60381.	3064.6
Stddev	33.3	69.	2.8
%RSD	.33435	.11363	.09242
#1	10010.	60351.	3061.5
#2	9961.5	60333.	3067.1
#3	9946.1	60460.	3065.2

Approved: June 27, 2016

Sample Name: L1606114801 Acquired: 6/25/2016 16:27:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.03462	.00065	.00821	.00587	.00014	1.4096	-.00013
Stddev	.00034	.00628	.00229	.00098	.00178	.00003	.0346	.00033
%RSD	57.407	18.141	351.58	11.960	30.235	19.203	2.4561	248.47
#1	.00054	.03990	-.00071	.00924	.00790	.00011	1.3926	.00011
#2	.00095	.02767	-.00063	.00729	.00463	.00014	1.4495	-.00000
#3	.00028	.03630	.00329	.00809	.00508	.00016	1.3868	-.00051
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.00001	.00030	.04215	.21418	.00024	.69121	.00161
Stddev	.00012	.00038	.00134	.02269	.07335	.00510	.07378	.00189
%RSD	127.75	2953.4	439.05	53.849	34.250	2116.0	10.674	117.34
#1	.00006	.00044	.00117	.03668	.16871	.00256	.67687	-.00053
#2	-.00001	-.00026	.00097	.06707	.29880	.00377	.77111	.00305
#3	.00022	-.00014	-.00123	.02268	.17502	-.00561	.62565	.00230
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00000	3.4091	-.00039	.01019	-.00019	-.00195	.00226	.25600
Stddev	.00027	.0299	.00017	.00142	.00263	.00116	.00241	.00788
%RSD	8569.8	.87827	42.602	13.923	1405.5	59.228	106.91	3.0774
#1	.00029	3.3757	-.00048	.01168	-.00018	-.00094	.00166	.24819
#2	-.00005	3.4183	-.00020	.01005	.00244	-.00171	.00492	.25587
#3	-.00025	3.4335	-.00049	.00885	-.00282	-.00321	.00020	.26394
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606114801 Acquired: 6/25/2016 16:27:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00043**	**.01395**	**.00070**	**-.00087**	**.00132**	**.00137**	**.00142**
Stddev	.00051	.00044	.00486	.00154	.00026	.00005	.00170
%RSD	118.61	3.1779	695.33	177.20	19.933	3.3203	119.70
#1	-.00063	.01431	.00581	-.00155	.00155	.00135	.00002
#2	.00015	.01345	.00013	-.00194	.00103	.00135	.00093
#3	-.00082	.01408	-.00385	.00089	.00136	.00143	.00331
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9913.2**	**60132.**	**3023.4**
Stddev	14.6	83.	23.3
%RSD	.14679	.13818	.77035
#1	9929.6	60151.	3030.8
#2	9901.6	60041.	2997.3
#3	9908.5	60204.	3042.1

Approved: June 27, 2016

Sample Name: L1606114802 Acquired: 6/25/2016 16:31:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00158	-.00408	.00026	.00633	.00904	.00015	2.7703	-.00011
Stddev	.00108	.00521	.00166	.00204	.00121	.00009	.0227	.00014
%RSD	68.438	127.59	628.67	32.297	13.379	59.313	.82092	123.10
#1	.00072	-.00923	-.00155	.00429	.00969	.00008	2.7476	-.00002
#2	.00123	.00119	.00169	.00631	.00978	.00025	2.7930	-.00027
#3	.00279	-.00421	.00065	.00838	.00764	.00012	2.7703	-.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00034	.00002	.00060	.02383	.06717	.00040	2.3757	.00083
Stddev	.00028	.00081	.00057	.01240	.09428	.00275	.0633	.00197
%RSD	81.635	5221.9	93.832	52.028	140.36	693.38	2.6650	237.49
#1	.00044	-.00078	.00046	.02298	.16159	-.00259	2.3743	.00224
#2	.00003	-.00000	.00012	.01188	-.02697	.00095	2.4397	.00167
#3	.00055	.00084	.00122	.03663	.06688	.00283	2.3131	-.00142
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00011	6.0259	-.00040	.00438	.00227	-.00140	.00185	.32995
Stddev	.00024	.0927	.00045	.00883	.00074	.00370	.00550	.00769
%RSD	208.31	1.5391	111.52	201.54	32.728	264.56	298.09	2.3316
#1	.00039	5.9438	-.00070	.01456	.00143	-.00567	.00483	.32126
#2	-.00003	6.1265	.00011	-.00011	.00282	.00058	.00521	.33270
#3	-.00002	6.0074	-.00061	-.00130	.00258	.00089	-.00450	.33589
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606114802 Acquired: 6/25/2016 16:31:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00030	.01996	.00356	-.00001	.00116	.00095	-.00024
Stddev	.00040	.00029	.00718	.00102	.00055	.00026	.00029
%RSD	136.64	1.4402	201.93	7537.5	47.304	27.165	121.82
#1	-.00006	.02008	.00861	-.00117	.00177	.00075	-.00049
#2	.00022	.02016	-.00466	.00037	.00099	.00086	-.00028
#3	.00074	.01963	.00672	.00076	.00071	.00124	.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9925.6	60099.	3031.8
Stddev	7.7	64.	18.1
%RSD	.07799	.10696	.59652
#1	9922.2	60109.	3022.9
#2	9920.2	60157.	3019.9
#3	9934.5	60030.	3052.6

Approved: June 27, 2016

Sample Name: L1606115001 Acquired: 6/25/2016 16:35:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00166	.05641	.00050	.00663	.00676	.00016	1.7818	.00015
Stddev	.00049	.00787	.00304	.00311	.00055	.00011	.0479	.00031
%RSD	29.631	13.947	601.42	46.927	8.1582	71.613	2.6886	203.48
#1	.00200	.06510	-.00294	.00991	.00739	.00005	1.7526	.00038
#2	.00109	.05435	.00166	.00372	.00635	.00014	1.8371	-.00020
#3	.00188	.04978	.00279	.00626	.00655	.00027	1.7557	.00027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00007	.00048	.00100	.04970	.00137	.00112	1.1341	.00143
Stddev	.00049	.00051	.00115	.01548	.09645	.00218	.0897	.00096
%RSD	745.24	106.17	114.47	31.142	7038.5	195.00	7.9106	67.253
#1	-.00002	.00105	.00223	.06756	-.03266	.00315	1.0361	.00043
#2	-.00057	.00034	.00081	.04145	.11022	-.00118	1.2123	.00235
#3	.00040	.00005	-.00004	.04010	-.07346	.00137	1.1537	.00152
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00027	5.9717	-.00015	.00950	-.00176	-.00302	.00622	.28058
Stddev	.00014	.0718	.00043	.00093	.00081	.00221	.00277	.00537
%RSD	53.804	1.2031	276.46	9.7789	45.909	73.167	44.558	1.9140
#1	-.00010	5.9074	-.00014	.01003	-.00249	-.00320	.00355	.27441
#2	-.00036	6.0493	-.00058	.01005	-.00089	-.00072	.00603	.28316
#3	-.00034	5.9584	.00027	.00843	-.00189	-.00513	.00908	.28418
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115001 Acquired: 6/25/2016 16:35:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00008	.01287	.00136	-.00194	.00110	.00114	.00062
Stddev	.00058	.00017	.00295	.00426	.00072	.00013	.00087
%RSD	762.02	1.3177	217.51	219.83	65.867	11.722	140.33
#1	.00015	.01301	.00162	-.00450	.00182	.00119	.00034
#2	-.00054	.01292	.00417	-.00429	.00110	.00099	-.00008
#3	.00061	.01268	-.00172	.00298	.00038	.00124	.00159
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9919.3	60059.	3016.8
Stddev	34.5	56.	13.7
%RSD	.34767	.09285	.45386
#1	9942.2	60062.	3013.8
#2	9936.0	60114.	3004.8
#3	9879.6	60002.	3031.7

Approved: June 27, 2016

Sample Name: L1606115102 Acquired: 6/25/2016 16:39:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00105	.00214	.00018	.00362	.00526	.00013	1.7820	.00006
Stddev	.00166	.00346	.00084	.00176	.00021	.00005	.0144	.00011
%RSD	158.53	161.68	476.67	48.690	4.0116	41.814	.80931	182.53
#1	.00213	.00451	-.00003	.00223	.00549	.00009	1.7780	.00003
#2	-.00086	-.00183	-.00054	.00560	.00521	.00010	1.7700	-.00003
#3	.00188	.00375	.00110	.00302	.00507	.00019	1.7980	.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00024	.00030	.00081	.02568	-.04054	.00233	1.0068	.00278
Stddev	.00009	.00056	.00053	.02210	.04687	.00374	.0618	.00296
%RSD	39.196	184.13	65.165	86.089	115.61	160.62	6.1428	106.56
#1	.00032	.00005	.00046	.05079	-.00224	.00526	1.0696	.00587
#2	.00013	-.00008	.00055	.00916	-.02657	.00361	.94591	-.00003
#3	.00026	.00094	.00141	.01708	-.09281	-.00188	1.0048	.00250
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00000	4.8272	.00042	.00764	.00031	-.00602	.00252	.26699
Stddev	.00027	.0498	.00058	.00427	.00107	.00140	.00418	.00420
%RSD	7088.8	1.0326	136.65	55.913	346.71	23.202	166.24	1.5735
#1	-.00018	4.7713	.00031	.01044	.00076	-.00742	-.00226	.26233
#2	-.00015	4.8672	-.00009	.00272	.00108	-.00462	.00551	.26815
#3	.00031	4.8430	.00105	.00974	-.00091	-.00602	.00431	.27049
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115102 Acquired: 6/25/2016 16:39:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00036	.01171	-.00105	-.00167	.00048	.00111	-.00010
Stddev	.00036	.00032	.00211	.00218	.00034	.00004	.00053
%RSD	98.761	2.7498	200.92	130.51	71.305	3.3481	503.34
#1	-.00064	.01138	-.00295	-.00411	.00028	.00107	-.00046
#2	-.00048	.01174	.00122	-.00099	.00029	.00113	.00050
#3	.00004	.01202	-.00143	.00009	.00088	.00113	-.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9924.7	60209.	3035.9
Stddev	21.0	164.	13.4
%RSD	.21110	.27319	.44278
#1	9945.2	60212.	3050.8
#2	9925.5	60043.	3024.7
#3	9903.3	60372.	3032.1

Approved: June 27, 2016

Sample Name: L1606115105 Acquired: 6/25/2016 16:43:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00211**	**.00219**	**.00055**	**.00586**	**.00711**	**.00018**	**2.2105**	**.00001**
Stddev	.00101	.00795	.00092	.00071	.00025	.00010	.0174	.00025
%RSD	47.839	363.29	167.06	12.119	3.4464	53.942	.78551	4197.7
#1	.00096	-.00697	.00162	.00634	.00685	.00013	2.1935	-.00020
#2	.00251	.00621	-.00001	.00619	.00715	.00012	2.2282	.00029
#3	.00286	.00732	.00005	.00504	.00733	.00030	2.2098	-.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00003**	**-.00001**	**.00002**	**.02737**	**.01189**	**.00109**	**1.3118**	**.00125**
Stddev	.00029	.00051	.00146	.02301	.05921	.00366	.0639	.00239
%RSD	827.06	7244.7	6104.0	84.080	497.84	335.35	4.8727	191.19
#1	.00035	.00040	.00134	.00788	.02968	-.00236	1.3142	.00227
#2	-.00003	.00015	.00027	.02147	.06017	.00070	1.3745	.00297
#3	-.00022	-.00058	-.00154	.05276	-.05417	.00493	1.2467	-.00148
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00023**	**6.1621**	**.00034**	**.00603**	**.00215**	**-.00121**	**.00114**	**.31741**
Stddev	.00059	.0852	.00146	.00590	.00148	.00098	.00444	.00674
%RSD	252.32	1.3824	433.95	97.933	68.608	80.696	390.40	2.1227
#1	-.00086	6.0786	-.00133	.00455	.00384	-.00161	.00564	.31056
#2	-.00015	6.2489	.00098	.00101	.00153	-.00010	.00100	.31765
#3	.00031	6.1588	.00136	.01253	.00108	-.00192	-.00323	.32403
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115105 Acquired: 6/25/2016 16:43:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00016**	**.01487**	**-.00053**	**-.00166**	**.00081**	**.00090**	**.00078**
Stddev	.00083	.00056	.00515	.00250	.00073	.00004	.00134
%RSD	533.32	3.7472	973.66	150.50	89.693	4.1812	172.14
#1	-.00110	.01433	.00470	-.00448	.00103	.00087	.00219
#2	.00015	.01545	-.00069	-.00074	.00000	.00094	.00064
#3	.00048	.01482	-.00560	.00025	.00141	.00089	-.00049
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9886.5**	**59788.**	**3007.6**
Stddev	13.0	185.	33.5
%RSD	.13157	.30931	1.1137
#1	9901.4	60001.	3037.0
#2	9880.8	59677.	2971.1
#3	9877.4	59685.	3014.8

Approved: June 27, 2016

Sample Name: L1606115401 Acquired: 6/25/2016 16:47:06 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573707-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00038	.00056	-.00040	.08474	.00202	.00013	25.077	-.00002
Stddev	.00048	.01146	.00100	.00137	.00093	.00006	.183	.00004
%RSD	125.93	2038.5	249.32	1.6179	46.237	47.434	.73077	179.25
#1	-.00014	.00766	-.00156	.08632	.00308	.00006	25.288	-.00005
#2	.00079	.00669	.00017	.08409	.00131	.00017	24.966	-.00003
#3	.00048	-.01266	.00019	.08382	.00167	.00016	24.976	.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00025	-.00049	.00090	.03824	.13932	.00586	5.7156	.01284
Stddev	.00012	.00049	.00033	.01141	.04539	.00340	.0899	.00260
%RSD	49.754	99.993	36.739	29.841	32.583	57.965	1.5726	20.218
#1	.00011	-.00085	.00123	.04902	.09367	.00206	5.7030	.01056
#2	.00028	-.00069	.00089	.02629	.18446	.00861	5.6327	.01230
#3	.00036	.00007	.00057	.03940	.13984	.00691	5.8112	.01567
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	48.626	-.00075	.01117	.00114	-.00491	.00349	.27480
Stddev	.00026	.111	.00066	.00745	.00133	.00248	.00272	.00312
%RSD	52.573	.22800	87.964	66.704	116.26	50.491	77.811	1.1358
#1	.00024	48.499	-.00062	.01513	.00176	-.00250	.00652	.27222
#2	.00075	48.707	-.00147	.01582	.00206	-.00745	.00269	.27392
#3	.00048	48.670	-.00016	.00258	-.00038	-.00478	.00127	.27827
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115401 Acquired: 6/25/2016 16:47:06 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573707-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.45176	-.00366	.00143	.00149	.00078	.00089
Stddev	.00028	.00268	.00362	.00047	.00058	.00004	.00106
%RSD	304.19	.59263	98.732	33.094	39.124	5.2122	119.26
#1	.00040	.44867	-.00000	.00088	.00209	.00083	-.00013
#2	-.00012	.45326	-.00375	.00168	.00093	.00076	.00081
#3	-.00001	.45335	-.00723	.00172	.00145	.00075	.00200
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9797.1	59243.	3012.9
Stddev	17.5	107.	17.7
%RSD	.17826	.18036	.58631
#1	9776.9	59361.	2997.1
#2	9806.9	59214.	3009.7
#3	9807.4	59153.	3032.0

Approved: June 27, 2016

Sample Name: L1606115402S Acquired: 6/25/2016 16:51:01 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: ~~WG573994-~~04 WG573707-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.01309**	**.25824**	**.01110**	**.13590**	**.02825**	**.00141**	**24.615**	**.00126**
Stddev	.00068	.00760	.00152	.00169	.00035	.00008	.204	.00013
%RSD	5.1924	2.9441	13.682	1.2425	1.2428	5.6800	.82995	10.655
#1	.01329	.24951	.01277	.13422	.02816	.00148	24.549	.00132
#2	.01365	.26341	.01072	.13589	.02864	.00133	24.844	.00111
#3	.01234	.26179	.00981	.13760	.02795	.00143	24.452	.00136
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00516**	**.01347**	**.01556**	**.12663**	**1.3534**	**.03179**	**5.7750**	**.02544**
Stddev	.00029	.00054	.00043	.02761	.0228	.00410	.0699	.00285
%RSD	5.6307	4.0382	2.7758	21.808	1.6862	12.889	1.2101	11.189
#1	.00539	.01407	.01514	.12414	1.3303	.03536	5.8428	.02774
#2	.00525	.01301	.01552	.15541	1.3541	.02732	5.7789	.02226
#3	.00483	.01334	.01600	.10034	1.3759	.03267	5.7032	.02633
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02479**	**48.072**	**.01332**	**.27819**	**.01575**	**.03066**	**.01661**	**.41681**
Stddev	.00011	.553	.00070	.01153	.00513	.00307	.00921	.00803
%RSD	.43243	1.1513	5.2238	4.1456	32.553	10.027	55.466	1.9271
#1	.02491	47.545	.01404	.27355	.01630	.02789	.02173	.40846
#2	.02475	48.649	.01266	.26969	.02058	.03397	.02212	.41747
#3	.02470	48.023	.01326	.29132	.01037	.03013	.00597	.42449
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115402S Acquired: 6/25/2016 16:51:01 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: ~~WG573994-04~~ WG573707-07

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02366**	**.45706**	**.02085**	**.01353**	**.02750**	**.02826**	**.00274**
Stddev	.00062	.00423	.00301	.00109	.00044	.00035	.00077
%RSD	2.6070	.92537	14.446	8.0746	1.5967	1.2498	28.151
#1	.02338	.45248	.01904	.01478	.02755	.02786	.00198
#2	.02436	.46082	.01917	.01306	.02704	.02846	.00352
#3	.02323	.45789	.02432	.01275	.02792	.02848	.00271
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9814.2**	**59172.**	**3008.6**
Stddev	14.0	133.	22.5
%RSD	.14277	.22528	.74717
#1	9829.8	59323.	3030.6
#2	9810.3	59073.	2985.6
#3	9802.6	59119.	3009.6

Approved: June 27, 2016

Sample Name: L1606115403SD Acquired: 6/25/2016 16:54:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: ~~WG573994-05~~ WG573707-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01239	.28748	.01145	.14150	.02764	.00148	24.958	.00130
Stddev	.00114	.00888	.00065	.00093	.00147	.00009	.077	.00038
%RSD	9.2124	3.0872	5.6507	.65421	5.3298	6.1781	.30772	29.306
#1	.01125	.28330	.01179	.14053	.02888	.00144	25.006	.00087
#2	.01353	.28146	.01070	.14159	.02802	.00141	24.999	.00161
#3	.01238	.29767	.01185	.14238	.02601	.00158	24.870	.00142
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00611	.01431	.01527	.13086	1.3857	.03280	5.7817	.02287
Stddev	.00054	.00136	.00145	.01260	.0370	.00239	.0561	.00299
%RSD	8.8509	9.4861	9.4751	9.6283	2.6697	7.2716	.97114	13.075
#1	.00671	.01585	.01436	.14435	1.3780	.03556	5.7473	.02433
#2	.00566	.01329	.01452	.11939	1.4259	.03138	5.8465	.01943
#3	.00597	.01379	.01694	.12885	1.3532	.03147	5.7513	.02484
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02737	48.881	.01504	.29126	.01658	.02795	.01955	.43672
Stddev	.00021	.600	.00152	.00741	.00249	.00243	.00943	.00693
%RSD	.75532	1.2279	10.076	2.5426	15.013	8.6768	48.242	1.5861
#1	.02720	48.195	.01331	.29382	.01763	.03060	.01049	.42891
#2	.02732	49.311	.01567	.29705	.01373	.02742	.02931	.43914
#3	.02760	49.136	.01613	.28292	.01836	.02584	.01885	.44211
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115403SD Acquired: 6/25/2016 16:54:54 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: ~~WG573994-05~~ WG573707-08

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02590**	**.47018**	**.02569**	**.01379**	**.02893**	**.02996**	**.00113**
Stddev	.00080	.00475	.00439	.00029	.00110	.00076	.00027
%RSD	3.0868	1.0106	17.097	2.0910	3.8168	2.5433	23.745
#1	.02499	.46471	.02819	.01349	.02776	.02918	.00138
#2	.02647	.47252	.02061	.01406	.02908	.03000	.00115
#3	.02625	.47331	.02825	.01382	.02996	.03071	.00085
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9727.4**	**58906.**	**3022.7**
Stddev	40.6	23.	25.7
%RSD	.41767	.03969	.84864
#1	9765.9	58911.	3039.9
#2	9731.2	58926.	2993.2
#3	9685.0	58880.	3035.0

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 16:58:48 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41361**	**10.445**	**.40516**	**.50612**	**1.0259**	**.05224**	**10.203**
Stddev	.00260	.021	.00144	.00287	.0033	.00020	.046
%RSD	.62922	.20256	.35469	.56757	.31739	.38953	.44884
#1	.41655	10.462	.40524	.50908	1.0271	.05242	10.224
#2	.41159	10.421	.40656	.50592	1.0284	.05202	10.235
#3	.41269	10.451	.40369	.50335	1.0222	.05229	10.150
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05215**	**.20471**	**.51698**	**.52477**	**4.0648**	**49.423**	**1.0320**
Stddev	.00013	.00021	.00148	.00103	.0369	.189	.0006
%RSD	.23994	.10391	.28587	.19687	.90838	.38306	.06170
#1	.05201	.20470	.51640	.52461	4.1007	49.428	1.0316
#2	.05218	.20450	.51865	.52588	4.0269	49.610	1.0327
#3	.05226	.20492	.51587	.52383	4.0668	49.231	1.0316
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.8590**	**.51551**	**1.0404**	**50.331**	**.51945**	**10.229**	**.51430**
Stddev	.0965	.00144	.0039	.173	.00106	.010	.00570
%RSD	.97927	.27876	.37864	.34422	.20472	.09385	1.1081
#1	9.9705	.51433	1.0396	50.428	.51822	10.232	.51032
#2	9.8060	.51711	1.0369	50.435	.52013	10.218	.51175
#3	9.8006	.51508	1.0447	50.131	.51999	10.236	.52083
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 16:58:48 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.2288	.41352	5.2624	1.0315	1.0269	1.0279	.53017
Stddev	.0104	.00168	.0015	.0026	.0016	.0023	.00514
%RSD	.84208	.40618	.02811	.25583	.15527	.22255	.96993
#1	1.2288	.41198	5.2628	1.0285	1.0283	1.0298	.53564
#2	1.2391	.41326	5.2608	1.0336	1.0273	1.0286	.52543
#3	1.2184	.41531	5.2637	1.0323	1.0252	1.0254	.52943
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	1.0564	1.0308	F 3.0180
Stddev	.0015	.0001	.0119
%RSD	.14240	.01349	.39369
#1	1.0580	1.0307	3.0121
#2	1.0561	1.0310	3.0317
#3	1.0550	1.0308	3.0102
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9528.1	57486.	2960.8
Stddev	9.3	201.	17.3
%RSD	.09776	.34979	.58425
#1	9517.6	57273.	2962.9
#2	9535.3	57514.	2942.6
#3	9531.5	57672.	2977.0

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 17:02:25 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00045**	**.00226**	**.00067**	**.00055**	**.00185**	**.00013**	**-.07167**	**-.00014**
Stddev	.00085	.00279	.00119	.00264	.00042	.00005	.02175	.00033
%RSD	189.88	123.24	177.00	483.17	22.573	35.869	30.346	234.73
#1	.00019	-.00053	.00152	-.00019	.00233	.00018	-.09322	-.00050
#2	-.00141	.00505	-.00069	-.00165	.00153	.00012	-.04973	.00014
#3	-.00012	.00226	.00119	.00348	.00171	.00009	-.07206	-.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00002**	**.00057**	**.00076**	**.03056**	**-.09613**	**.00228**	**-.13259**	**.00051**
Stddev	.00039	.00056	.00130	.01338	.02053	.00471	.02062	.00136
%RSD	1720.9	98.019	170.60	43.800	21.361	206.77	15.553	268.18
#1	.00020	.00017	-.00074	.01806	-.09726	-.00131	-.14855	-.00092
#2	.00020	.00121	.00148	.02892	-.07505	.00054	-.10931	.00179
#3	-.00047	.00034	.00155	.04468	-.11607	.00761	-.13993	.00066
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00325**	**.04047**	**-.00081**	**.00150**	**.00148**	**.00530**	**.00256**	**-.00673**
Stddev	.00022	.01175	.00096	.00183	.00289	.00260	.00237	.00167
%RSD	6.7467	29.033	119.00	121.89	195.62	48.929	92.639	24.783
#1	.00325	.05342	-.00016	.00278	.00470	.00570	.00309	-.00596
#2	.00303	.03049	-.00036	.00231	.00063	.00768	-.00003	-.00558
#3	.00347	.03751	-.00192	-.00059	-.00090	.00253	.00461	-.00864
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 17:02:25 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00041**	**.00044**	**.00426**	**-.00209**	**.00094**	**.00001**	**.00353**
Stddev	.00052	.00045	.00463	.00321	.00096	.00021	.00070
%RSD	128.76	102.93	108.80	153.83	102.68	1611.6	19.702
#1	.00048	.00069	.00959	-.00418	.00024	-.00012	.00320
#2	.00089	.00071	.00120	.00161	.00204	-.00009	.00306
#3	-.00015	-.00008	.00198	-.00369	.00053	.00025	.00433
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9729.2**	**58924.**	**2982.2**
Stddev	14.1	124.	20.8
%RSD	.14444	.21107	.69802
#1	9725.0	58912.	2962.2
#2	9744.9	58806.	3003.8
#3	9717.8	59054.	2980.7

Approved: June 27, 2016

Sample Name: L1606115501 Acquired: 6/25/2016 17:06:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00033**	**.00180**	**.00125**	**.01161**	**.01032**	**.00011**	**2.3136**	**-.00007**
Stddev	.00230	.00043	.00160	.00137	.00099	.00007	.0269	.00019
%RSD	694.54	23.757	128.16	11.776	9.5715	61.393	1.1630	281.96
#1	.00225	.00218	.00194	.01272	.00925	.00003	2.3397	.00009
#2	-.00214	.00134	.00239	.01008	.01120	.00016	2.2860	-.00001
#3	-.00111	.00188	-.00058	.01203	.01051	.00015	2.3151	-.00027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**.00014**	**-.00018**	**.02935**	**.09569**	**-.00047**	**1.4383**	**.08572**
Stddev	.00040	.00048	.00052	.00252	.08702	.00314	.0635	.00277
%RSD	831.42	341.03	289.17	8.5796	90.931	664.03	4.4136	3.2299
#1	.00005	.00069	-.00054	.02667	.18563	.00276	1.4626	.08626
#2	.00029	-.00017	.00042	.03166	.01192	-.00066	1.4861	.08273
#3	-.00048	-.00010	-.00042	.02972	.08953	-.00352	1.3663	.08818
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00023**	**3.8750**	**.00064**	**.00454**	**.00264**	**-.00168**	**.00429**	**.37174**
Stddev	.00028	.0249	.00026	.00324	.00293	.00204	.00272	.00091
%RSD	124.52	.64140	39.830	71.430	111.04	121.37	63.452	.24386
#1	-.00008	3.8959	.00063	.00081	.00037	-.00400	.00736	.37278
#2	.00028	3.8475	.00040	.00665	.00594	-.00087	.00338	.37129
#3	.00048	3.8816	.00091	.00616	.00160	-.00017	.00214	.37115
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115501 Acquired: 6/25/2016 17:06:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00004**	**.01613**	**.00436**	**.00049**	**-.00004**	**.00087**	**.00142**
Stddev	.00092	.00051	.00120	.00199	.00066	.00012	.00191
%RSD	2230.2	3.1791	27.456	410.68	1766.6	13.339	134.89
#1	.00087	.01607	.00571	.00178	-.00027	.00083	.00247
#2	-.00002	.01565	.00341	-.00181	.00071	.00100	-.00079
#3	-.00097	.01667	.00397	.00149	-.00055	.00078	.00257
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9938.9**	**60110.**	**3041.6**
Stddev	14.3	99.	8.4
%RSD	.14384	.16440	.27454
#1	9922.4	60057.	3050.1
#2	9946.9	60224.	3033.5
#3	9947.4	60049.	3041.2

Approved: June 27, 2016

Sample Name: L1606115502 Acquired: 6/25/2016 17:10:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00066	-.00453	-.00033	.00989	.01794	.00013	2.4889	.00011
Stddev	.00144	.01600	.00099	.00264	.00078	.00009	.0280	.00014
%RSD	216.56	353.18	295.78	26.673	4.3403	71.235	1.1261	129.84
#1	.00231	.00655	-.00001	.00702	.01871	.00024	2.4937	.00027
#2	.00001	-.02287	-.00144	.01220	.01715	.00010	2.5143	-.00000
#3	-.00033	.00273	.00045	.01047	.01798	.00006	2.4588	.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00008	-.00044	.00104	.24605	-.07520	-.00217	1.6643	.02363
Stddev	.00022	.00012	.00019	.01179	.12175	.00286	.0624	.00224
%RSD	272.66	27.286	18.718	4.7938	161.89	131.42	3.7522	9.4751
#1	.00027	-.00041	.00090	.23638	.00559	-.00035	1.5996	.02553
#2	.00015	-.00033	.00126	.24256	-.01597	-.00070	1.6689	.02116
#3	-.00017	-.00057	.00095	.25919	-.21524	-.00546	1.7243	.02420
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00021	2.8989	.00016	.00365	.00260	-.00217	.00870	.42450
Stddev	.00065	.0101	.00120	.00396	.00147	.00150	.00507	.00601
%RSD	314.77	.35003	759.93	108.45	56.477	68.876	58.304	1.4160
#1	-.00082	2.8946	-.00099	.00772	.00422	-.00382	.01102	.41775
#2	-.00027	2.9105	.00140	-.00019	.00221	-.00091	.01219	.42650
#3	.00047	2.8917	.00006	.00342	.00137	-.00178	.00288	.42926
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115502 Acquired: 6/25/2016 17:10:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00038**	**.02139**	**.00454**	**-.00050**	**.00021**	**.00050**	**.00118**
Stddev	.00016	.00022	.00843	.00114	.00027	.00017	.00137
%RSD	41.986	1.0329	185.67	229.51	129.67	34.277	116.06
#1	.00030	.02116	.00016	.00076	.00041	.00055	-.00040
#2	.00027	.02139	.01426	-.00146	.00031	.00031	.00208
#3	.00056	.02161	-.00080	-.00078	-.00010	.00064	.00187
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9945.8**	**60248.**	**3055.3**
Stddev	9.2	94.	11.4
%RSD	.09217	.15541	.37444
#1	9945.9	60314.	3061.2
#2	9954.9	60289.	3042.1
#3	9936.5	60141.	3062.6

Approved: June 27, 2016

Sample Name: L1606115503 Acquired: 6/25/2016 17:14:07 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00200	-.00135	-.00104	.00757	.00737	.00014	2.3997	-.00010
Stddev	.00192	.00661	.00046	.00116	.00061	.00003	.0309	.00022
%RSD	96.141	489.89	44.488	15.290	8.2167	21.217	1.2897	218.06
#1	.00406	-.00575	-.00051	.00624	.00716	.00012	2.3659	-.00020
#2	.00025	-.00456	-.00131	.00835	.00689	.00017	2.4268	-.00025
#3	.00170	.00625	-.00131	.00813	.00805	.00012	2.4063	.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00001	-.00044	-.00019	.10432	-.04481	.00004	.91519	.02191
Stddev	.00035	.00082	.00101	.01415	.09851	.00279	.09371	.00121
%RSD	3296.1	188.98	541.26	13.565	219.83	6329.0	10.240	5.5295
#1	-.00033	.00027	.00027	.10406	-.13456	.00163	.99135	.02326
#2	-.00007	-.00023	-.00134	.09030	-.06048	.00168	.94369	.02093
#3	.00037	-.00134	.00052	.11860	.06059	-.00317	.81054	.02153
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00009	2.4900	.00017	.00956	.00115	-.00111	.00305	.32330
Stddev	.00005	.0408	.00057	.00296	.00162	.00041	.00349	.00479
%RSD	56.574	1.6398	345.24	30.912	140.85	37.279	114.64	1.4819
#1	-.00011	2.4429	.00078	.00752	.00187	-.00066	-.00055	.31801
#2	-.00013	2.5161	.00008	.00822	-.00070	-.00119	.00642	.32454
#3	-.00003	2.5109	-.00036	.01295	.00229	-.00148	.00327	.32735
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606115503 Acquired: 6/25/2016 17:14:07 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00009**	**.01387**	**.00305**	**-.00065**	**.00052**	**.00084**	**.00071**
Stddev	.00081	.00016	.00789	.00174	.00073	.00004	.00140
%RSD	953.07	1.1448	258.94	265.86	139.13	5.0666	198.28
#1	-.00013	.01403	.00198	-.00023	.00032	.00082	.00133
#2	-.00088	.01386	-.00426	.00083	-.00008	.00082	.00169
#3	.00075	.01371	.01142	-.00257	.00133	.00089	-.00090
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9910.7**	**60184.**	**3019.4**
Stddev	10.3	175.	9.0
%RSD	.10381	.29035	.29953
#1	9910.7	60334.	3026.1
#2	9921.0	60225.	3023.1
#3	9900.4	59992.	3009.1

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 17:18:04 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41531**	**10.447**	**.40916**	**.50533**	**1.0238**	**.05217**	**10.225**
Stddev	.00091	.032	.00132	.00054	.0034	.00011	.030
%RSD	.21998	.31066	.32348	.10618	.32870	.20293	.29201
#1	.41613	10.474	.41061	.50566	1.0229	.05229	10.191
#2	.41432	10.411	.40886	.50561	1.0276	.05213	10.247
#3	.41546	10.455	.40801	.50471	1.0211	.05209	10.235
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05225**	**.20370**	**.51712**	**.52549**	**4.0444**	**49.392**	**1.0301**
Stddev	.00031	.00034	.00175	.00197	.0184	.117	.0048
%RSD	.59442	.16603	.33927	.37570	.45604	.23752	.46659
#1	.05229	.20396	.51906	.52418	4.0394	49.465	1.0330
#2	.05254	.20382	.51666	.52776	4.0648	49.455	1.0327
#3	.05192	.20332	.51564	.52452	4.0290	49.257	1.0245
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.7951**	**.51327**	**1.0423**	**50.236**	**.51912**	**10.208**	**.51997**
Stddev	.0472	.00504	.0084	.169	.00170	.017	.00372
%RSD	.48165	.98200	.80990	.33677	.32826	.16636	.71476
#1	9.7523	.51223	1.0511	50.262	.52046	10.219	.52035
#2	9.7875	.51875	1.0343	50.391	.51968	10.189	.51607
#3	9.8457	.50883	1.0415	50.056	.51720	10.217	.52348
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 17:18:04 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2238**	**.40709**	**5.2468**	**1.0260**	**1.0230**	**1.0304**	**.52875**
Stddev	.0084	.00511	.0098	.0012	.0023	.0127	.00327
%RSD	.68659	1.2545	.18586	.11540	.22459	1.2311	.61859
#1	1.2260	.40344	5.2580	1.0271	1.0231	1.0165	.53252
#2	1.2308	.40491	5.2428	1.0262	1.0252	1.0415	.52673
#3	1.2145	.41292	5.2398	1.0248	1.0206	1.0331	.52699
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0570**	**1.0310**	F **3.0075**
Stddev	.0056	.0001	.0207
%RSD	.53103	.00939	.68856
#1	1.0633	1.0311	3.0004
#2	1.0553	1.0309	3.0308
#3	1.0525	1.0310	2.9913
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9557.2**	**57623.**	**2979.5**
Stddev	9.7	193.	16.3
%RSD	.10109	.33465	.54600
#1	9546.7	57405.	2979.8
#2	9565.7	57771.	2963.0
#3	9559.4	57693.	2995.6

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 17:21:42 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**-.00563**	**.00200**	**-.00100**	**.00189**	**.00009**	**-.01127**	**-.00004**
Stddev	.00104	.00334	.00233	.00184	.00120	.00004	.03123	.00010
%RSD	1246.5	59.297	116.54	184.11	63.245	40.940	277.20	226.69
#1	-.00074	-.00335	.00466	-.00138	.00302	.00011	-.03377	.00004
#2	-.00027	-.00407	.00108	-.00263	.00202	.00005	.02439	-.00015
#3	.00126	-.00946	.00027	.00100	.00064	.00010	-.02442	-.00003
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00016**	**.00048**	**.00089**	**.00955**	**-.03243**	**.00071**	**-.01124**	**-.00013**
Stddev	.00017	.00075	.00043	.01226	.03763	.00268	.07316	.00129
%RSD	103.61	157.05	48.720	128.38	116.03	378.57	650.98	975.30
#1	.00002	.00095	.00059	-.00410	-.06003	-.00175	-.02665	-.00039
#2	.00035	.00087	.00069	.01962	.01043	.00031	-.07547	-.00128
#3	.00012	-.00039	.00138	.01314	-.04771	.00356	.06840	.00127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00323**	**.02898**	**-.00046**	**.00432**	**.00328**	**.00454**	**.00410**	**-.00425**
Stddev	.00029	.02008	.00095	.00359	.00221	.00123	.00133	.00064
%RSD	9.1168	69.292	208.78	82.910	67.490	26.995	32.528	15.135
#1	.00298	.02490	-.00038	.00845	.00352	.00323	.00275	-.00470
#2	.00356	.01126	-.00144	.00261	.00095	.00566	.00541	-.00351
#3	.00316	.05079	.00046	.00192	.00536	.00474	.00413	-.00454
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 17:21:42 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00068	.00021	.00375	-.00110	.00108	-.00009	.00393
Stddev	.00059	.00011	.00446	.00333	.00027	.00003	.00087
%RSD	86.694	51.296	119.05	302.86	25.045	36.058	22.188
#1	.00135	.00012	.00567	.00135	.00136	-.00006	.00491
#2	.00024	.00018	-.00135	.00024	.00107	-.00012	.00363
#3	.00044	.00033	.00692	-.00489	.00082	-.00008	.00324
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9706.7	59138.	3007.6
Stddev	9.6	122.	14.5
%RSD	.09901	.20648	.48341
#1	9697.1	58997.	2992.2
#2	9716.3	59217.	3009.4
#3	9706.6	59200.	3021.1

Approved: June 27, 2016

Sample Name: LLCCV Acquired: 6/25/2016 17:25:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00682	.16326	.00787	.08110	.01015	.00172	.37504	.00080
Stddev	.00230	.01176	.00393	.00058	.00060	.00002	.03274	.00014
%RSD	33.789	7.2031	49.909	.72089	5.8834	1.2548	8.7301	17.457
#1	.00456	.16635	.01239	.08173	.01084	.00173	.33805	.00070
#2	.00916	.17317	.00587	.08058	.00985	.00173	.40031	.00075
#3	.00674	.15026	.00534	.08099	.00976	.00169	.38675	.00096
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00446	.00390	.00484	.09497	.71355	.08241	.33888	.00741
Stddev	.00036	.00087	.00066	.01610	.00812	.00379	.00870	.00176
%RSD	8.1660	22.291	13.622	16.951	1.1379	4.5949	2.5673	23.707
#1	.00488	.00490	.00494	.08463	.70937	.08497	.33660	.00920
#2	.00426	.00332	.00413	.08675	.70838	.08419	.34850	.00734
#3	.00424	.00348	.00544	.11352	.72291	.07806	.33155	.00569
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00855	.40833	.01678	.78916	.01199	.08269	.02089	.77549
Stddev	.00007	.02661	.00076	.00972	.00016	.00111	.00270	.00482
%RSD	.82014	6.5175	4.5074	1.2317	1.3733	1.3390	12.938	.62171
#1	.00856	.42291	.01686	.80033	.01215	.08397	.01985	.78057
#2	.00847	.42447	.01749	.78261	.01182	.08216	.01886	.77494
#3	.00861	.37762	.01598	.78454	.01200	.08195	.02396	.77097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LLCCV Acquired: 6/25/2016 17:25:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40481**	**.04214**	**.02257**	**.15413**	**.00994**	**.01699**	**.13476**
Stddev	.00262	.00003	.00404	.00144	.00108	.00007	.00396
%RSD	.64734	.07891	17.895	.93373	10.832	.40613	2.9378
#1	.40763	.04217	.02259	.15502	.01117	.01705	.13268
#2	.40435	.04210	.02661	.15491	.00942	.01692	.13226
#3	.40245	.04214	.01853	.15247	.00922	.01700	.13932
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9757.6**	**59331.**	**3007.6**
Stddev	3.7	156.	34.2
%RSD	.03744	.26375	1.1357
#1	9754.3	59400.	3008.1
#2	9757.0	59152.	3041.5
#3	9761.5	59441.	2973.2

Approved: June 27, 2016

Sample Name: PBW Y5 Acquired: 6/25/2016 17:29:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00185**	**.00116**	**-.00112**	**-.00084**	**.00120**	**.00011**	**-.01961**	**-.00006**
Stddev	.00045	.01041	.00198	.00240	.00080	.00010	.05200	.00016
%RSD	24.231	897.41	176.31	284.10	67.234	89.198	265.17	288.76
#1	.00153	-.00604	-.00020	.00070	.00046	.00020	.01383	-.00024
#2	.00236	-.00358	.00023	-.00360	.00107	.00014	-.07952	.00007
#3	.00166	.01310	-.00339	.00037	.00206	.00000	.00686	.00000
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00010**	**-.00018**	**.00105**	**.03412**	**-.07407**	**.00236**	**.02694**	**.00162**
Stddev	.00024	.00064	.00070	.02542	.08098	.00211	.02742	.00224
%RSD	248.33	362.92	67.168	74.514	109.32	89.477	101.81	137.94
#1	.00030	-.00083	.00024	.00498	-.16192	.00417	.01237	.00092
#2	.00016	-.00014	.00146	.05181	-.05788	.00285	.00987	.00413
#3	-.00017	.00044	.00145	.04557	-.00242	.00004	.05857	-.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00014**	**.01025**	**-.00077**	**.00507**	**.00386**	**-.00192**	**-.00109**	**.00437**
Stddev	.00004	.01479	.00185	.00580	.00053	.00084	.00245	.00068
%RSD	28.058	144.27	239.74	114.27	13.817	43.798	224.68	15.497
#1	-.00010	.02026	-.00256	.00797	.00348	-.00273	.00112	.00408
#2	-.00015	-.00674	.00113	.00885	.00447	-.00198	-.00067	.00389
#3	-.00017	.01724	-.00088	-.00160	.00363	-.00105	-.00372	.00515
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: PBW Y5 Acquired: 6/25/2016 17:29:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00077	.00044	.00538	-.00184	.00071	.00062	.00198
Stddev	.00057	.00040	.00350	.00250	.00039	.00016	.00115
%RSD	73.146	89.773	65.128	135.94	54.397	26.345	58.223
#1	.00143	.00035	.00942	-.00427	.00062	.00079	.00279
#2	.00040	.00088	.00331	.00073	.00114	.00048	.00066
#3	.00049	.00010	.00340	-.00198	.00038	.00058	.00249
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9880.3	60109.	3027.1
Stddev	31.5	73.	6.6
%RSD	.31885	.12179	.21761
#1	9847.3	60149.	3030.6
#2	9910.1	60154.	3019.5
#3	9883.5	60025.	3031.2

Approved: June 27, 2016

Sample Name: LCSW Y5 Acquired: 6/25/2016 17:33:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.18967	4.7620	.18828	.90013	.47555	.02395	4.7950	.02373
Stddev	.00250	.0081	.00104	.00471	.00273	.00006	.0213	.00018
%RSD	1.3192	.17078	.55148	.52284	.57347	.24848	.44480	.74408
#1	.18907	4.7578	.18718	.89893	.47827	.02388	4.7819	.02392
#2	.19242	4.7714	.18841	.90532	.47282	.02399	4.8196	.02370
#3	.18753	4.7569	.18924	.89614	.47557	.02397	4.7835	.02357
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09489	.23803	.24292	1.9123	22.817	.48875	4.5669	.24181
Stddev	.00039	.00120	.00156	.0298	.036	.00284	.0465	.00007
%RSD	.41346	.50457	.64104	1.5597	.15629	.58110	1.0173	.02751
#1	.09520	.23794	.24291	1.9317	22.780	.48603	4.5750	.24188
#2	.09501	.23927	.24447	1.8779	22.852	.49170	4.5170	.24174
#3	.09445	.23687	.24136	1.9272	22.818	.48851	4.6088	.24180
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.49027	23.490	.24091	4.5553	.24063	.56646	.18382	2.3525
Stddev	.00145	.113	.00040	.0084	.00352	.00190	.00239	.0014
%RSD	.29550	.48187	.16805	.18553	1.4632	.33509	1.2995	.06062
#1	.48963	23.608	.24065	4.5528	.24014	.56531	.18109	2.3541
#2	.49193	23.382	.24137	4.5647	.24437	.56542	.18555	2.3519
#3	.48925	23.480	.24069	4.5484	.23738	.56865	.18482	2.3515
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LCSW Y5 Acquired: 6/25/2016 17:33:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47085**	**.48454**	**.47923**	**.24546**	**.48968**	**.47202**	**.01198**
Stddev	.00082	.00070	.00116	.00117	.00078	.00042	.00247
%RSD	.17390	.14504	.24256	.47677	.15993	.08855	20.639
#1	.47170	.48509	.47789	.24652	.48877	.47178	.00945
#2	.47079	.48375	.47984	.24421	.49011	.47251	.01208
#3	.47006	.48479	.47995	.24566	.49015	.47178	.01439
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9888.6**	**60001.**	**3090.1**
Stddev	19.1	149.	26.5
%RSD	.19299	.24892	.85667
#1	9894.9	60052.	3063.1
#2	9867.1	59834.	3091.3
#3	9903.7	60119.	3116.0

Approved: June 27, 2016

Sample Name: F BLANK Acquired: 6/25/2016 17:37:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00166	-.00148	.00090	.00227	.00163	.00011	-.00468	.00008
Stddev	.00130	.00490	.00188	.00068	.00037	.00003	.02546	.00022
%RSD	78.305	330.55	208.29	29.974	22.968	25.204	544.18	269.24
#1	.00033	-.00239	.00275	.00286	.00197	.00009	-.03398	.00015
#2	.00173	.00381	-.00100	.00243	.00123	.00014	.01206	.00026
#3	.00293	-.00586	.00096	.00153	.00167	.00010	.00788	-.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00010	.00058	.00129	.04476	.01060	-.00161	-.05710	.00063
Stddev	.00029	.00122	.00029	.02153	.02075	.00622	.06303	.00318
%RSD	276.09	211.55	22.793	48.099	195.81	387.03	110.39	502.59
#1	-.00039	.00184	.00121	.03412	-.00619	.00062	-.09553	.00029
#2	.00018	-.00059	.00161	.06953	.03381	.00319	-.09140	.00397
#3	-.00010	.00048	.00104	.03061	.00418	-.00863	.01564	-.00236
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00056	.05677	-.00011	.00316	.00576	-.00381	.00560	-.00010
Stddev	.00040	.03222	.00050	.00162	.00079	.00329	.00231	.00141
%RSD	72.047	56.761	453.37	51.247	13.705	86.399	41.256	1388.8
#1	.00085	.07680	-.00069	.00292	.00611	-.00013	.00434	-.00092
#2	.00073	.01960	.00014	.00168	.00631	-.00483	.00420	-.00091
#3	.00010	.07391	.00021	.00489	.00485	-.00647	.00827	.00153
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: F BLANK Acquired: 6/25/2016 17:37:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00007	.00021	-.00396	-.00081	.00112	.00164	.00166
Stddev	.00084	.00011	.00377	.00288	.00053	.00013	.00011
%RSD	1212.1	52.839	95.183	357.63	47.087	8.1028	6.9152
#1	-.00082	.00025	-.00813	-.00413	.00144	.00151	.00159
#2	.00085	.00029	-.00300	.00065	.00141	.00165	.00179
#3	.00018	.00008	-.00077	.00106	.00051	.00177	.00159
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10015.	61314.	3090.8
Stddev	16.	177.	23.0
%RSD	.15522	.28891	.74279
#1	10012.	61143.	3064.6
#2	10032.	61303.	3107.1
#3	10001.	61497.	3100.8

Approved: June 27, 2016

Sample Name: L1606098602 Acquired: 6/25/2016 17:41:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00088	-.00042	-.00040	.09099	.02315	.00010	31.081	-.00006
Stddev	.00141	.00402	.00168	.00026	.00060	.00002	.292	.00002
%RSD	160.55	951.46	416.45	.28845	2.5949	20.817	.93935	37.551
#1	.00197	.00339	-.00222	.09093	.02306	.00008	31.173	-.00006
#2	-.00072	-.00463	.00109	.09127	.02259	.00012	30.754	-.00004
#3	.00139	-.00003	-.00008	.09076	.02379	.00010	31.315	-.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00024	.00059	.00094	.04555	.98402	.01199	5.7014	.03480
Stddev	.00048	.00016	.00041	.01573	.08850	.00357	.0718	.00212
%RSD	197.55	26.299	43.232	34.544	8.9935	29.756	1.2595	6.0810
#1	.00072	.00042	.00054	.02747	.88485	.01459	5.6202	.03718
#2	-.00024	.00064	.00135	.05303	1.0122	.01345	5.7566	.03314
#3	.00026	.00072	.00092	.05614	1.0550	.00792	5.7274	.03406
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00142	119.95	.00080	.00488	.00205	-.00278	-.00166	3.2347
Stddev	.00007	.78	.00132	.00386	.00179	.00227	.00673	.0109
%RSD	4.5853	.65032	165.26	79.044	87.096	81.844	406.14	.33815
#1	.00137	120.20	.00201	.00062	.00137	-.00037	-.00009	3.2461
#2	.00149	119.07	.00100	.00813	.00409	-.00489	.00415	3.2336
#3	.00140	120.57	-.00061	.00590	.00071	-.00307	-.00903	3.2243
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098602 Acquired: 6/25/2016 17:41:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00169	.54147	-.00161	-.00445	.00062	.00187	.00166
Stddev	.00060	.00457	.00418	.00188	.00027	.00011	.00091
%RSD	35.360	.84355	259.35	42.200	43.191	5.8148	54.635
#1	.00106	.54395	-.00040	-.00269	.00035	.00198	.00205
#2	.00225	.53620	.00182	-.00642	.00063	.00186	.00062
#3	.00175	.54425	-.00626	-.00422	.00089	.00176	.00230
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9793.0	59457.	3063.6
Stddev	12.2	310.	25.9
%RSD	.12488	.52191	.84399
#1	9783.4	59142.	3057.4
#2	9788.8	59467.	3092.0
#3	9806.7	59762.	3041.4

Approved: June 27, 2016

Sample Name: L1606098604 Acquired: 6/25/2016 17:45:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00188**	**.00192**	**.00240**	**.01933**	**.01606**	**.00017**	**258.27**
Stddev	.00168	.00663	.00025	.00142	.00064	.00002	.74
%RSD	89.608	344.23	10.527	7.3208	4.0034	12.550	.28551
#1	.00265	-.00052	.00267	.01828	.01569	.00019	258.08
#2	.00303	-.00313	.00217	.01877	.01568	.00016	257.65
#3	-.00005	.00943	.00237	.02094	.01680	.00015	259.09
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00045**	**.00038**	**.00085**	**.01097**	**.02808**	**1.7243**	**.01739**
Stddev	.00023	.00042	.00027	.00107	.01426	.0758	.00077
%RSD	50.915	110.30	31.401	9.7605	50.795	4.3949	4.4280
#1	.00057	.00083	.00065	.01199	.02723	1.7078	.01656
#2	.00019	.00033	.00115	.00986	.01426	1.6581	.01808
#3	.00061	-.00001	.00075	.01107	.04275	1.8069	.01753
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**67.436**	**1.3917**	**.00099**	**14.380**	**.00387**	**.00116**	**.00177**
Stddev	.211	.0043	.00004	.031	.00051	.00485	.00143
%RSD	.31235	.30797	4.3492	.21337	13.178	419.65	80.619
#1	67.195	1.3899	.00099	14.411	.00414	.00253	.00287
#2	67.533	1.3886	.00103	14.350	.00328	-.00423	.00229
#3	67.581	1.3966	.00095	14.377	.00419	.00517	.00016
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: L1606098604 Acquired: 6/25/2016 17:45:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00130**	**.00146**	**3.7576**	**.00044**	**1.2471**	F **-.03982**	**-.00203**
Stddev	.00495	.00434	.0110	.00062	.0034	.00406	.00082
%RSD	381.66	297.39	.29276	142.20	.26857	10.186	40.503
#1	-.00325	.00313	3.7635	.00113	1.2483	-.04298	-.00122
#2	.00433	.00472	3.7644	.00025	1.2433	-.03524	-.00200
#3	-.00498	-.00347	3.7449	-.00007	1.2496	-.04122	-.00286
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00050**	**.03051**	**.00151**
Stddev	.00029	.00016	.00074
%RSD	57.554	.53953	48.929
#1	-.00084	.03044	.00164
#2	-.00033	.03070	.00216
#3	-.00034	.03040	.00071
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9610.5**	**58897.**	**3044.6**
Stddev	3.8	97.	5.6
%RSD	.03917	.16551	.18408
#1	9614.6	58944.	3043.4
#2	9607.2	58961.	3050.7
#3	9609.7	58785.	3039.7

Approved: June 27, 2016

Sample Name: L1606098606 Acquired: 6/25/2016 17:48:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00223	.00972	-.00038	.01373	.06613	.00026	77.440	.00018
Stddev	.00105	.00309	.00132	.00205	.00021	.00009	.075	.00021
%RSD	47.129	31.833	346.72	14.955	.32162	33.087	.09643	116.74
#1	.00235	.00691	-.00086	.01602	.06623	.00027	77.417	.00042
#2	.00112	.00920	-.00139	.01205	.06589	.00034	77.379	.00001
#3	.00321	.01303	.00111	.01311	.06628	.00017	77.523	.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00039	.00108	.01311	.03469	1.0666	.00528	22.814	.00226
Stddev	.00015	.00032	.00053	.01962	.0266	.00092	.288	.00095
%RSD	37.987	29.419	4.0055	56.546	2.4929	17.478	1.2638	41.753
#1	.00056	.00082	.01330	.03776	1.0701	.00541	22.650	.00309
#2	.00027	.00143	.01252	.05259	1.0384	.00430	22.645	.00123
#3	.00036	.00097	.01352	.01372	1.0912	.00613	23.147	.00247
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00038	16.230	.00086	.01073	.00099	-.00405	.00160	2.3863
Stddev	.00023	.056	.00046	.00069	.00181	.00317	.00143	.0169
%RSD	59.834	.34372	53.276	6.4188	182.38	78.406	89.367	.70593
#1	.00032	16.197	.00073	.01029	.00246	-.00129	-.00004	2.4028
#2	.00063	16.198	.00137	.01153	-.00103	-.00752	.00258	2.3871
#3	.00019	16.294	.00048	.01038	.00155	-.00333	.00227	2.3691
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098606 Acquired: 6/25/2016 17:48:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00004	.39726	-.00928	-.00135	-.00044	.03812	.00202
Stddev	.00081	.00077	.00327	.00251	.00044	.00033	.00118
%RSD	1967.6	.19375	35.277	185.96	101.24	.85896	58.324
#1	.00088	.39645	-.00551	-.00111	-.00048	.03844	.00108
#2	-.00002	.39799	-.01103	.00103	.00002	.03814	.00334
#3	-.00074	.39735	-.01131	-.00397	-.00086	.03779	.00164
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9875.4	60112.	3085.9
Stddev	17.7	60.	4.4
%RSD	.17969	.09972	.14239
#1	9863.6	60084.	3084.9
#2	9866.7	60181.	3090.7
#3	9895.8	60072.	3082.1

Approved: June 27, 2016

Sample Name: L1606098608 Acquired: 6/25/2016 17:52:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00010	.00115	.00027	.01585	.03262	.00011	84.217	.00013
Stddev	.00235	.00513	.00111	.00136	.00141	.00004	.314	.00017
%RSD	2475.5	446.86	409.02	8.5842	4.3250	37.715	.37277	125.68
#1	-.00262	.00404	-.00099	.01740	.03376	.00010	84.484	.00019
#2	.00142	-.00477	.00110	.01531	.03104	.00016	84.297	.00026
#3	.00149	.00417	.00071	.01485	.03305	.00008	83.871	-.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00015	.00177	.00567	.03434	1.2401	.01415	23.228	.02727
Stddev	.00046	.00068	.00058	.01787	.0425	.00095	.227	.00106
%RSD	299.90	38.444	10.195	52.028	3.4256	6.6956	.97712	3.8827
#1	-.00049	.00241	.00506	.05413	1.2665	.01352	23.122	.02850
#2	.00037	.00184	.00573	.01939	1.1911	.01369	23.489	.02663
#3	-.00034	.00105	.00621	.02950	1.2628	.01524	23.073	.02669
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00074	12.660	.00033	.00413	.00179	-.00334	-.00002	3.4869
Stddev	.00017	.055	.00044	.00392	.00093	.00378	.00639	.0225
%RSD	23.632	.43231	134.67	94.801	51.837	113.14	36088.	.64396
#1	.00093	12.716	.00084	.00852	.00134	-.00400	-.00400	3.4968
#2	.00069	12.657	.00002	.00099	.00117	-.00675	-.00341	3.5027
#3	.00059	12.607	.00013	.00288	.00285	.00073	.00736	3.4612
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098608 Acquired: 6/25/2016 17:52:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00019**	**.83876**	**-.00998**	**-.00331**	**.00001**	**.01681**	**.00058**
Stddev	.00022	.00442	.00442	.00242	.00077	.00018	.00259
%RSD	112.09	.52651	44.282	73.168	6824.2	1.0650	446.63
#1	-.00039	.84240	-.01091	-.00183	-.00011	.01673	-.00103
#2	-.00022	.84004	-.00517	-.00610	-.00069	.01701	.00357
#3	.00004	.83385	-.01387	-.00199	.00083	.01668	-.00080
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9863.7**	**59966.**	**3069.4**
Stddev	33.8	107.	16.4
%RSD	.34299	.17793	.53396
#1	9879.3	59844.	3052.8
#2	9824.9	60042.	3069.7
#3	9886.9	60012.	3085.6

Approved: June 27, 2016

Sample Name: L1606098608PS Acquired: 6/25/2016 17:56:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG574216-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.19012	4.7962	.18985	.92761	.50579	.02416	80.589	.02376
Stddev	.00115	.0181	.00324	.00188	.00153	.00010	.509	.00031
%RSD	.60416	.37710	1.7085	.20237	.30247	.42369	.63122	1.3080
#1	.18881	4.7997	.18690	.92863	.50534	.02415	80.494	.02389
#2	.19096	4.8123	.19332	.92876	.50750	.02427	81.138	.02341
#3	.19058	4.7766	.18933	.92544	.50454	.02406	80.134	.02399
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09205	.23624	.24131	1.9065	24.291	.50517	25.434	.26103
Stddev	.00018	.00146	.00130	.0143	.039	.00496	.150	.00626
%RSD	.19260	.61641	.53797	.74911	.16017	.98165	.59175	2.3977
#1	.09210	.23659	.24071	1.9177	24.287	.49968	25.563	.26590
#2	.09186	.23464	.24280	1.9113	24.254	.50647	25.268	.26320
#3	.09221	.23749	.24041	1.8904	24.332	.50934	25.470	.25397
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.49345	34.822	.23345	4.6884	.23609	.56850	.18268	5.5114
Stddev	.00019	.169	.00142	.0142	.00258	.00539	.00604	.0369
%RSD	.03860	.48533	.60979	.30359	1.0940	.94754	3.3047	.66920
#1	.49323	34.790	.23475	4.6782	.23789	.57472	.18356	5.4689
#2	.49357	35.005	.23367	4.6823	.23725	.56528	.18824	5.5341
#3	.49354	34.671	.23193	4.7047	.23313	.56551	.17626	5.5313
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098608PS Acquired: 6/25/2016 17:56:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG574216-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47094**	**1.2232**	**.47334**	**.23988**	**.49331**	**.47631**	**.00522**
Stddev	.00189	.0075	.00962	.00117	.00125	.00347	.00156
%RSD	.40190	.61543	2.0323	.48785	.25347	.72879	29.825
#1	.47068	1.2227	.46319	.24082	.49476	.47230	.00351
#2	.47295	1.2310	.48233	.23857	.49254	.47818	.00558
#3	.46920	1.2160	.47451	.24025	.49265	.47845	.00657
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9831.9**	**59526.**	**3075.8**
Stddev	15.2	194.	19.0
%RSD	.15421	.32611	.61679
#1	9829.8	59604.	3066.0
#2	9817.9	59305.	3063.8
#3	9848.0	59669.	3097.7

Approved: June 27, 2016

Sample Name: L1606098608SDL Acquired: 6/25/2016 18:00:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG574216-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00260**	**-.00514**	**.00187**	**.00508**	**.00868**	**.00012**	**17.605**	**-.00015**
Stddev	.00156	.01146	.00424	.00200	.00063	.00008	.142	.00010
%RSD	60.035	222.87	226.36	39.430	7.2120	66.284	.80739	66.159
#1	.00437	.00780	.00118	.00739	.00923	.00013	17.451	-.00025
#2	.00143	-.00924	.00642	.00396	.00881	.00020	17.731	-.00007
#3	.00199	-.01399	-.00198	.00389	.00800	.00004	17.633	-.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00018**	**.00006**	**.00193**	**.02399**	**.21059**	**-.00038**	**4.9921**	**.00509**
Stddev	.00017	.00044	.00078	.00583	.02318	.00216	.0797	.00111
%RSD	97.856	772.57	40.443	24.318	11.008	572.98	1.5970	21.792
#1	-.00004	-.00041	.00282	.01859	.23319	.00193	5.0036	.00623
#2	-.00037	.00046	.00138	.03017	.18686	-.00072	4.9072	.00402
#3	-.00011	.00012	.00158	.02319	.21171	-.00234	5.0654	.00500
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00034**	**2.6677**	**-.00071**	**.00513**	**.00009**	**-.00286**	**-.00058**	**.72991**
Stddev	.00017	.0114	.00035	.00552	.00151	.00279	.00302	.00273
%RSD	51.097	.42653	49.400	107.62	1685.9	97.489	522.92	.37384
#1	.00017	2.6710	-.00044	.00122	.00180	.00001	.00149	.72939
#2	.00033	2.6550	-.00058	.01144	-.00102	-.00304	-.00405	.73285
#3	.00051	2.6770	-.00110	.00273	-.00051	-.00555	.00083	.72747
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098608SDL Acquired: 6/25/2016 18:00:25 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG574216-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00006	.17684	.00226	.00019	.00052	.00413	.00011
Stddev	.00027	.00102	.00232	.00368	.00057	.00014	.00068
%RSD	464.39	.57572	102.44	1951.0	110.67	3.3962	623.50
#1	-.00013	.17748	.00403	.00128	.00008	.00397	.00059
#2	.00024	.17567	-.00036	.00320	.00031	.00420	.00041
#3	-.00028	.17737	.00312	-.00392	.00117	.00423	-.00067
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9896.0	60261.	3042.3
Stddev	25.0	81.	28.3
%RSD	.25242	.13391	.92921
#1	9907.7	60289.	3040.8
#2	9913.0	60324.	3071.3
#3	9867.3	60170.	3014.9

Approved: June 27, 2016

Sample Name: L1606098608SDL Acquired: 6/25/2016 18:04:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG574216-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00175	-.00807	-.00063	-.00078	.00182	.00014	3.3399	-.00019
Stddev	.00147	.00382	.00127	.00273	.00016	.00003	.0946	.00023
%RSD	83.969	47.279	200.74	348.35	8.5330	23.588	2.8330	121.57
#1	.00097	-.00874	.00010	.00038	.00196	.00018	3.2721	-.00045
#2	.00084	-.00396	.00010	.00117	.00166	.00011	3.4480	-.00012
#3	.00344	-.01151	-.00210	-.00390	.00185	.00014	3.2995	-.00000
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00003	-.00037	.00107	.01769	-.08284	.00119	.94569	.00052
Stddev	.00047	.00133	.00130	.01702	.09737	.00358	.05932	.00055
%RSD	1445.4	360.33	120.72	96.212	117.53	302.26	6.2728	105.20
#1	.00035	-.00086	.00167	-.00174	-.18470	.00532	.88043	.00057
#2	-.00050	-.00138	-.00041	.02995	-.07314	-.00085	.96028	.00105
#3	.00025	.00113	.00196	.02486	.00931	-.00092	.99635	-.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00012	.49987	-.00021	.00755	.00303	-.00541	.00850	.13812
Stddev	.00032	.04592	.00020	.00386	.00188	.00314	.00116	.00966
%RSD	265.16	9.1856	96.762	51.054	62.049	58.070	13.660	6.9939
#1	.00001	.44687	-.00020	.01173	.00086	-.00531	.00838	.12921
#2	-.00049	.52761	-.00041	.00680	.00410	-.00232	.00740	.13676
#3	.00011	.52513	-.00001	.00413	.00413	-.00859	.00972	.14839
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098608SDL Acquired: 6/25/2016 18:04:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG574216-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00039**	**.03350**	**.00349**	**.00073**	**.00046**	**.00106**	**.00161**
Stddev	.00032	.00105	.00672	.00069	.00093	.00001	.00112
%RSD	80.289	3.1425	192.33	93.894	202.19	.93440	69.701
#1	.00063	.03282	.00553	-.00006	-.00031	.00105	.00058
#2	.00051	.03471	.00896	.00114	.00020	.00107	.00146
#3	.00004	.03296	-.00401	.00113	.00149	.00107	.00281
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10400.**	**61068.**	**3085.8**
Stddev	302.	416.	38.3
%RSD	2.9005	.68044	1.2414
#1	10744.	61402.	3085.7
#2	10277.	61200.	3047.5
#3	10180.	60603.	3124.1

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 18:08:15 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.40600	10.196	.39897	.49353	1.0097	.05111	10.072
Stddev	.00134	.005	.00384	.00145	.0055	.00016	.108
%RSD	.32964	.04609	.96241	.29468	.54041	.30777	1.0683
#1	.40454	10.197	.40224	.49289	1.0058	.05115	10.014
#2	.40717	10.200	.39474	.49519	1.0073	.05125	10.006
#3	.40630	10.191	.39992	.49250	1.0159	.05094	10.196
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05085	.19930	.50343	.51248	4.0210	48.769	1.0170
Stddev	.00016	.00069	.00173	.00042	.0444	.141	.0095
%RSD	.31221	.34558	.34304	.08145	1.1049	.29004	.93218
#1	.05072	.20005	.50491	.51259	3.9923	48.842	1.0084
#2	.05079	.19915	.50386	.51282	3.9985	48.605	1.0154
#3	.05103	.19869	.50153	.51201	4.0722	48.858	1.0271
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.6470	.50568	1.0217	49.589	.50657	9.9463	.50897
Stddev	.0505	.00379	.0050	.266	.00174	.0145	.00224
%RSD	.52402	.74931	.49273	.53683	.34444	.14541	.44016
#1	9.7040	.50521	1.0275	49.574	.50855	9.9602	.50639
#2	9.6294	.50215	1.0196	49.330	.50589	9.9313	.51042
#3	9.6076	.50969	1.0181	49.862	.50527	9.9474	.51010
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 18:08:15 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1949**	**.39405**	**5.1402**	**1.0022**	**1.0107**	**.99866**	**.51518**
Stddev	.0031	.00179	.0100	.0005	.0053	.01659	.00058
%RSD	.25554	.45308	.19427	.05270	.52786	1.6608	.11222
#1	1.1981	.39492	5.1482	1.0028	1.0100	1.0024	.51475
#2	1.1944	.39525	5.1290	1.0020	1.0058	.98052	.51584
#3	1.1920	.39200	5.1434	1.0018	1.0164	1.0131	.51497
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0309**	**1.0070**	F **2.9451**
Stddev	.0008	.0014	.0223
%RSD	.08066	.13880	.75803
#1	1.0316	1.0084	2.9555
#2	1.0300	1.0069	2.9195
#3	1.0311	1.0056	2.9603
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9649.2**	**58096.**	**2982.6**
Stddev	5.7	114.	17.5
%RSD	.05915	.19560	.58802
#1	9648.4	58223.	2978.9
#2	9644.0	58003.	3001.7
#3	9655.3	58062.	2967.2

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 18:11:52 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00062**	**.00113**	**.00060**	**-.00134**	**.00269**	**.00016**	**-.00509**	**-.00011**
Stddev	.00447	.01287	.00139	.00141	.00058	.00006	.00999	.00008
%RSD	723.56	1142.0	234.30	105.23	21.456	39.000	196.19	75.399
#1	-.00262	.00580	.00212	.00028	.00233	.00022	-.01646	-.00020
#2	.00450	-.01343	-.00061	-.00230	.00238	.00017	.00228	-.00009
#3	-.00373	.01101	.00028	-.00199	.00335	.00009	-.00109	-.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00010**	**.00031**	**.00131**	**.01880**	**-.14052**	**-.00173**	**.03086**	**-.00243**
Stddev	.00043	.00144	.00103	.01556	.08453	.00290	.07623	.00189
%RSD	438.61	470.86	78.889	82.745	60.152	168.14	247.00	77.536
#1	.00038	.00014	.00051	.03667	-.04368	.00153	-.05463	-.00186
#2	-.00045	-.00104	.00248	.01148	-.17839	-.00266	.09175	-.00090
#3	-.00022	.00182	.00094	.00826	-.19949	-.00404	.05546	-.00454
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00375**	**-.00702**	**.00000**	**.00544**	**-.00002**	**.00180**	**.00158**	**.03715**
Stddev	.00019	.01848	.00030	.00537	.00129	.00320	.00255	.00232
%RSD	5.0585	263.40	8008.1	98.759	6600.5	177.29	161.59	6.2433
#1	.00382	-.02711	-.00030	.01149	.00142	.00145	-.00134	.03895
#2	.00354	-.00319	.00002	.00357	-.00108	.00516	.00266	.03453
#3	.00390	.00925	.00029	.00125	-.00040	-.00120	.00341	.03795
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 18:11:52 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00026	.00036	-.00017	-.00186	.00075	.00012	.00425
Stddev	.00040	.00044	.01020	.00315	.00028	.00007	.00095
%RSD	153.60	122.95	5879.7	169.59	36.989	59.021	22.345
#1	.00000	.00086	.00219	-.00527	.00044	.00006	.00508
#2	.00073	.00017	.00864	-.00126	.00081	.00010	.00321
#3	.00006	.00004	-.01135	.00095	.00098	.00019	.00446
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9372.1	56632.	2874.9
Stddev	24.8	50.	14.1
%RSD	.26420	.08878	.48885
#1	9350.8	56589.	2886.0
#2	9366.2	56619.	2859.1
#3	9399.3	56687.	2879.7

Approved: June 27, 2016

Sample Name: L1606098610 Acquired: 6/25/2016 18:15:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00168	.00613	.00026	.00622	.01341	.00005	115.37	.00029
Stddev	.00132	.00327	.00219	.00120	.00108	.00003	.56	.00010
%RSD	78.702	53.342	840.89	19.355	8.0675	55.551	.48696	35.866
#1	.00017	.00551	-.00222	.00662	.01310	.00009	115.81	.00019
#2	.00222	.00322	.00107	.00487	.01252	.00003	114.73	.00040
#3	.00265	.00967	.00193	.00717	.01462	.00004	115.55	.00029
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00004	.00098	.00166	.06083	.38326	.00831	27.726	.01242
Stddev	.00045	.00011	.00061	.01132	.08932	.00154	.200	.00158
%RSD	1035.7	11.515	36.634	18.614	23.306	18.467	.72137	12.708
#1	-.00034	.00088	.00210	.06356	.35599	.00912	27.699	.01068
#2	-.00027	.00110	.00191	.04839	.48304	.00927	27.542	.01377
#3	.00048	.00096	.00096	.07054	.31076	.00654	27.939	.01280
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00076	14.397	.00169	.00670	.00176	-.00418	.00480	3.9389
Stddev	.00003	.096	.00045	.00675	.00408	.00345	.00343	.0208
%RSD	3.3704	.66673	26.485	100.67	231.83	82.446	71.530	.52736
#1	.00075	14.373	.00205	.00909	-.00024	-.00395	.00094	3.9592
#2	.00074	14.315	.00119	.01193	.00646	-.00774	.00595	3.9398
#3	.00079	14.503	.00185	-.00091	-.00093	-.00086	.00752	3.9176
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098610 Acquired: 6/25/2016 18:15:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00034**	**.52737**	**-.01662**	**-.00476**	**.00053**	**.01128**	**.00653**
Stddev	.00061	.00238	.00515	.00157	.00069	.00008	.00193
%RSD	181.06	.45087	31.015	32.913	130.43	.73884	29.481
#1	-.00078	.52702	-.01240	-.00297	.00012	.01121	.00451
#2	.00036	.52519	-.01509	-.00547	.00132	.01137	.00674
#3	-.00059	.52991	-.02236	-.00585	.00014	.01126	.00835
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9795.7**	**59337.**	**3075.9**
Stddev	21.1	178.	25.1
%RSD	.21493	.29915	.81568
#1	9771.4	59262.	3048.2
#2	9807.4	59210.	3082.7
#3	9808.3	59540.	3097.0

Approved: June 27, 2016

Sample Name: L1606098612 Acquired: 6/25/2016 18:19:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00237	.00267	-.00060	.08716	.06662	.00041	142.66	.00007
Stddev	.00073	.00275	.00337	.00141	.00149	.00002	1.19	.00021
%RSD	30.688	102.87	562.71	1.6192	2.2393	3.9782	.83233	319.26
#1	.00315	.00518	-.00334	.08576	.06575	.00041	141.33	.00027
#2	.00225	-.00027	-.00161	.08858	.06577	.00039	143.01	.00007
#3	.00171	.00310	.00316	.08713	.06835	.00042	143.63	-.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00010	.00135	.04528	.02649	1.6602	.01963	40.720	.03253
Stddev	.00034	.00053	.00134	.03733	.1018	.00353	.262	.00324
%RSD	339.50	39.036	2.9680	140.93	6.1325	17.994	.64244	9.9559
#1	.00034	.00170	.04668	.04343	1.6128	.02218	40.440	.02879
#2	-.00029	.00162	.04400	-.01631	1.5907	.02112	40.763	.03448
#3	.00024	.00075	.04516	.05234	1.7770	.01560	40.958	.03432
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00115	26.540	.00106	.02312	.00172	-.00084	.00204	2.4275
Stddev	.00029	.172	.00094	.00593	.00284	.00147	.00269	.0046
%RSD	25.080	.64693	88.979	25.661	164.80	174.96	131.78	.18880
#1	.00089	26.345	.00153	.02349	.00164	-.00125	.00490	2.4289
#2	.00110	26.607	.00167	.01701	.00461	-.00205	-.00043	2.4311
#3	.00146	26.668	-.00003	.02886	-.00107	.00079	.00165	2.4223
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098612 Acquired: 6/25/2016 18:19:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00022**	**1.7410**	**-.01626**	**-.00241**	**.00038**	**.05602**	**.00415**
Stddev	.00061	.0138	.00313	.00137	.00051	.00009	.00053
%RSD	273.86	.79166	19.237	56.738	134.14	.16489	12.835
#1	-.00019	1.7255	-.01265	-.00306	.00050	.05591	.00443
#2	.00092	1.7457	-.01795	-.00084	.00081	.05608	.00354
#3	-.00006	1.7519	-.01818	-.00334	-.00018	.05607	.00449
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9747.5**	**58989.**	**3047.5**
Stddev	4.1	78.	13.8
%RSD	.04198	.13200	.45213
#1	9749.1	58936.	3063.4
#2	9750.6	59079.	3039.1
#3	9742.9	58952.	3040.0

Approved: June 27, 2016

Sample Name: L1606098614 Acquired: 6/25/2016 18:23:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00279	.01008	-.00066	.03390	.03124	.00010	196.20	.00021
Stddev	.00112	.00383	.00147	.00206	.00013	.00007	.85	.00008
%RSD	40.043	38.038	221.13	6.0906	.42986	73.782	.43251	39.178
#1	.00183	.00766	.00090	.03506	.03108	.00016	195.71	.00030
#2	.00402	.00807	-.00201	.03513	.03132	.00002	195.70	.00014
#3	.00253	.01449	-.00089	.03152	.03131	.00013	197.18	.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	.00088	.00840	.02428	2.9211	.02334	75.647	.04008
Stddev	.00023	.00045	.00064	.00605	.0291	.00193	.677	.00318
%RSD	446.71	50.770	7.6701	24.929	.99693	8.2676	.89445	7.9254
#1	-.00007	.00063	.00780	.03006	2.8958	.02251	74.950	.04023
#2	.00031	.00062	.00908	.02477	2.9529	.02197	75.691	.03683
#3	-.00009	.00140	.00832	.01799	2.9145	.02555	76.301	.04318
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00071	22.556	.00040	.01184	.00385	-.00016	.00418	3.4833
Stddev	.00033	.097	.00156	.00464	.00193	.00051	.00681	.0215
%RSD	46.348	.43032	390.32	39.236	50.052	315.18	162.82	.61768
#1	.00068	22.526	-.00110	.00655	.00167	-.00006	.01171	3.5066
#2	.00040	22.478	.00029	.01372	.00533	-.00072	-.00154	3.4792
#3	.00106	22.665	.00201	.01524	.00455	.00029	.00237	3.4641
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098614 Acquired: 6/25/2016 18:23:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**2.3670**	**-.02425**	**-.00232**	**-.00024**	**.02160**	**.00232**
Stddev	.00092	.0076	.00398	.00117	.00100	.00003	.00161
%RSD	715.81	.32195	16.401	50.456	415.47	.12320	69.325
#1	.00091	2.3659	-.02580	-.00236	.00091	.02157	.00413
#2	-.00082	2.3599	-.02722	-.00113	-.00075	.02159	.00107
#3	-.00048	2.3751	-.01973	-.00348	-.00089	.02163	.00175
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9640.5**	**58425.**	**3058.6**
Stddev	20.6	100.	4.7
%RSD	.21337	.17110	.15390
#1	9631.4	58473.	3059.7
#2	9664.1	58492.	3062.6
#3	9626.1	58310.	3053.4

Approved: June 27, 2016

Sample Name: L1606098616 Acquired: 6/25/2016 18:27:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00115	.03265	.00256	.11861	.01251	.00029	229.75	.00020
Stddev	.00195	.00899	.00533	.00161	.00061	.00014	.36	.00016
%RSD	169.27	27.532	208.27	1.3580	4.8819	47.504	.15530	78.536
#1	.00069	.04218	-.00197	.11885	.01280	.00014	229.65	.00038
#2	-.00053	.02433	.00844	.12008	.01292	.00030	229.45	.00006
#3	.00329	.03143	.00122	.11689	.01181	.00042	230.15	.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00062	.00178	.06744	.03108	3.2126	.03682	84.360	.29574
Stddev	.00023	.00069	.00047	.01537	.1307	.00386	.426	.00133
%RSD	37.410	38.666	.69994	49.454	4.0677	10.480	.50488	.44809
#1	.00042	.00192	.06781	.02007	3.0741	.03526	83.896	.29633
#2	.00087	.00239	.06691	.04863	3.2299	.03400	84.450	.29422
#3	.00056	.00104	.06760	.02452	3.3338	.04122	84.734	.29667
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00072	71.143	.00123	.03078	.00375	-.00200	.00018	5.1619
Stddev	.00027	.049	.00055	.00199	.00114	.00238	.00481	.0107
%RSD	37.642	.06818	44.575	6.4658	30.424	119.08	2715.0	.20775
#1	.00061	71.192	.00158	.03108	.00244	-.00009	.00517	5.1511
#2	.00102	71.095	.00060	.02865	.00425	-.00467	-.00442	5.1621
#3	.00051	71.141	.00151	.03260	.00456	-.00125	-.00023	5.1725
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098616 Acquired: 6/25/2016 18:27:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00037	6.5212	-.02963	-.00177	-.00109	.08946	.00332
Stddev	.00060	.0079	.00654	.00230	.00079	.00019	.00190
%RSD	161.30	.12151	22.085	130.49	72.258	.21177	57.202
#1	.00105	6.5204	-.02657	.00051	-.00123	.08936	.00540
#2	.00018	6.5138	-.02518	-.00171	-.00179	.08968	.00287
#3	-.00011	6.5296	-.03715	-.00409	-.00024	.08933	.00168
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9566.8	58078.	3050.1
Stddev	16.9	52.	19.7
%RSD	.17631	.08955	.64584
#1	9563.8	58018.	3068.4
#2	9585.0	58111.	3029.2
#3	9551.7	58106.	3052.7

Approved: June 27, 2016

Sample Name: L1606098618 Acquired: 6/25/2016 18:31:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00155	-.00096	.00289	.04251	.02915	.00005	62.293	-.00011
Stddev	.00038	.00238	.00029	.00109	.00082	.00008	.092	.00019
%RSD	24.410	246.77	9.9664	2.5737	2.7985	149.80	.14819	166.73
#1	.00113	.00168	.00257	.04189	.03009	-.00003	62.335	-.00017
#2	.00185	-.00164	.00314	.04186	.02874	.00005	62.357	.00010
#3	.00168	-.00294	.00296	.04377	.02862	.00013	62.187	-.00027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.00031	.00347	.01624	1.9071	.02198	20.259	.01093
Stddev	.00021	.00046	.00016	.00938	.1214	.00171	.067	.00376
%RSD	243.30	146.24	4.6588	57.772	6.3644	7.7782	.33311	34.347
#1	.00028	.00083	.00329	.00573	1.8306	.02394	20.336	.01303
#2	.00011	.00015	.00350	.01922	2.0470	.02118	20.230	.00660
#3	-.00013	-.00004	.00361	.02377	1.8435	.02082	20.211	.01318
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00102	25.999	-.00042	.00444	.00244	-.00359	.00392	3.1846
Stddev	.00023	.100	.00069	.00545	.00095	.00188	.00301	.0187
%RSD	22.864	.38558	161.73	122.87	39.121	52.433	76.809	.58628
#1	.00106	26.085	-.00101	-.00067	.00273	-.00311	.00526	3.2018
#2	.00123	26.024	.00033	.00381	.00321	-.00199	.00047	3.1874
#3	.00077	25.889	-.00059	.01018	.00137	-.00566	.00602	3.1647
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098618 Acquired: 6/25/2016 18:31:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00057	1.4545	-.00154	-.00233	.00037	.00832	.00208
Stddev	.00132	.0032	.00357	.00066	.00141	.00007	.00021
%RSD	228.95	.22347	232.21	28.221	386.86	.88012	10.206
#1	.00206	1.4546	-.00325	-.00241	-.00025	.00825	.00199
#2	-.00043	1.4577	-.00393	-.00294	-.00063	.00831	.00233
#3	.00009	1.4512	.00257	-.00163	.00198	.00839	.00193
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9841.2	59299.	3061.4
Stddev	105.5	272.	4.9
%RSD	1.0718	.45797	.16152
#1	9746.5	58986.	3061.4
#2	9822.2	59458.	3056.4
#3	9954.9	59454.	3066.3

Approved: June 27, 2016

Sample Name: L1606098620 Acquired: 6/25/2016 18:35:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00004	.03312	.00066	.07608	.03474	.00017	4.5650	.00010
Stddev	.00141	.00304	.00168	.00356	.00133	.00010	.0156	.00016
%RSD	3777.7	9.1736	255.45	4.6823	3.8419	59.446	.34129	166.50
#1	-.00004	.03663	-.00075	.08018	.03521	.00029	4.5712	.00011
#2	.00149	.03142	.00252	.07372	.03577	.00014	4.5766	.00025
#3	-.00133	.03132	.00020	.07433	.03323	.00009	4.5473	-.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00112	.00060	.00299	.05475	.56414	.00988	1.7714	.01450
Stddev	.00021	.00082	.00143	.01752	.07965	.00243	.0462	.00322
%RSD	18.384	137.46	47.864	31.992	14.119	24.596	2.6071	22.223
#1	.00132	.00110	.00244	.06300	.59505	.00894	1.8247	.01687
#2	.00115	.00105	.00462	.03463	.62370	.00806	1.7436	.01579
#3	.00091	-.00035	.00192	.06662	.47367	.01264	1.7459	.01083
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00137	251.01	-.00057	.03777	.00214	-.00589	.00108	2.0603
Stddev	.00029	.75	.00178	.00427	.00057	.00380	.00694	.0083
%RSD	21.156	.29997	313.28	11.296	26.364	64.542	640.75	.40056
#1	.00125	251.39	-.00025	.04217	.00279	-.00381	-.00651	2.0621
#2	.00170	251.50	-.00248	.03365	.00178	-.01028	.00710	2.0676
#3	.00115	250.15	.00103	.03749	.00186	-.00358	.00266	2.0514
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098620 Acquired: 6/25/2016 18:35:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00101**	**.14703**	**.00470**	**-.00231**	**.00100**	**.00923**	**.00423**
Stddev	.00046	.00057	.00299	.00179	.00041	.00019	.00094
%RSD	44.993	.38637	63.686	77.674	41.001	2.0938	22.219
#1	.00049	.14737	.00735	-.00044	.00130	.00917	.00391
#2	.00129	.14734	.00531	-.00246	.00116	.00906	.00349
#3	.00126	.14637	.00145	-.00402	.00053	.00944	.00529
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9725.3**	**58367.**	**3028.8**
Stddev	17.5	120.	15.3
%RSD	.18040	.20576	.50451
#1	9742.7	58420.	3020.7
#2	9707.6	58451.	3019.2
#3	9725.6	58230.	3046.4

Approved: June 27, 2016

Sample Name: L1606098622 Acquired: 6/25/2016 18:39:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00117	-.00596	.00037	.05747	.02764	.00023	127.72	.00011
Stddev	.00093	.00925	.00122	.00204	.00088	.00005	1.04	.00012
%RSD	79.226	155.22	327.64	3.5486	3.1759	23.660	.81279	108.09
#1	.00125	-.01464	.00026	.05716	.02813	.00022	127.35	.00021
#2	.00205	-.00701	-.00078	.05560	.02662	.00018	126.92	.00013
#3	.00020	.00377	.00164	.05965	.02815	.00028	128.89	-.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00027	.00090	.01009	.02031	2.5853	.01995	32.558	.04793
Stddev	.00056	.00109	.00042	.00894	.0945	.00518	.278	.00190
%RSD	208.67	121.92	4.1494	44.026	3.6562	25.950	.85453	3.9665
#1	.00085	.00146	.01042	.02811	2.6682	.02326	32.522	.05007
#2	.00021	-.00036	.01022	.02229	2.4824	.01399	32.299	.04730
#3	-.00026	.00160	.00962	.01055	2.6053	.02261	32.852	.04643
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00038	46.975	.00038	.00992	.00378	-.00381	.00198	3.5382
Stddev	.00001	.389	.00033	.00514	.00270	.00385	.00283	.0119
%RSD	1.5146	.82774	87.053	51.791	71.542	100.88	142.89	.33659
#1	.00038	46.842	.00000	.00464	.00569	-.00293	.00001	3.5245
#2	.00039	46.670	.00061	.01491	.00068	-.00048	.00522	3.5450
#3	.00039	47.413	.00052	.01022	.00497	-.00802	.00070	3.5451
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098622 Acquired: 6/25/2016 18:39:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00017	1.7827	-.01653	-.00317	-.00057	.00858	.00262
Stddev	.00056	.0144	.00634	.00502	.00088	.00017	.00024
%RSD	326.45	.80805	38.350	158.35	154.19	1.9458	9.0067
#1	-.00048	1.7747	-.02093	-.00538	-.00155	.00859	.00236
#2	.00051	1.7740	-.00927	.00258	.00014	.00841	.00281
#3	.00049	1.7993	-.01940	-.00671	-.00029	.00875	.00270
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9602.2	57749.	2959.9
Stddev	40.4	252.	26.3
%RSD	.42032	.43708	.88902
#1	9555.9	57561.	2977.4
#2	9630.3	57650.	2972.6
#3	9620.3	58036.	2929.6

Approved: June 27, 2016

Sample Name: L1606098624 Acquired: 6/25/2016 18:42:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00037**	**.00561**	**.00154**	**.40348**	**.02348**	**.00039**	**237.24**
Stddev	.00142	.00394	.00506	.00070	.00039	.00006	.43
%RSD	386.92	70.246	327.66	.17357	1.6661	16.452	.18001
#1	.00014	.00128	.00004	.40369	.02381	.00037	237.56
#2	-.00093	.00658	-.00259	.40270	.02305	.00046	237.41
#3	.00189	.00898	.00718	.40405	.02358	.00034	236.75
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**-.00007**	**.00152**	**.02649**	**.02448**	**5.9223**	**.01174**
Stddev	.00017	.00052	.00056	.00200	.00699	.0587	.00396
%RSD	362.68	725.31	36.507	7.5377	28.540	.99123	33.722
#1	-.00014	.00052	.00185	.02553	.03123	5.8563	.01398
#2	-.00015	-.00040	.00183	.02879	.02493	5.9416	.00717
#3	.00015	-.00034	.00088	.02516	.01728	5.9689	.01408
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**66.698**	**.32515**	**.00019**	**63.762**	**.00021**	**.01502**	**.00144**
Stddev	.426	.00095	.00039	.087	.00030	.00323	.00065
%RSD	.63897	.29073	207.41	.13653	145.82	21.500	44.732
#1	66.222	.32513	-.00023	63.794	-.00012	.01765	.00108
#2	67.045	.32421	.00055	63.829	.00049	.01601	.00219
#3	66.826	.32610	.00024	63.663	.00025	.01142	.00106
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: L1606098624 Acquired: 6/25/2016 18:42:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-02

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00149**	**.00522**	**3.3709**	**-.00072**	**1.6690**	F **-.03355**	**-.00155**
Stddev	.00099	.00177	.0249	.00122	.0015	.00264	.00229
%RSD	66.177	33.876	.73770	169.83	.08806	7.8818	147.53
#1	-.00080	.00419	3.3514	-.00109	1.6686	-.03207	.00107
#2	-.00262	.00726	3.3989	-.00171	1.6706	-.03199	-.00316
#3	-.00106	.00421	3.3624	.00064	1.6678	-.03661	-.00256
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00040**	**.00877**	**.00210**
Stddev	.00049	.00015	.00054
%RSD	123.14	1.7509	25.986
#1	-.00082	.00894	.00208
#2	.00014	.00865	.00156
#3	-.00050	.00870	.00265
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9491.7**	**57630.**	**3009.2**
Stddev	78.4	156.	15.6
%RSD	.82592	.27138	.51996
#1	9550.0	57451.	2991.3
#2	9402.6	57694.	3015.7
#3	9522.6	57743.	3020.5

Approved: June 27, 2016

Sample Name: L1606098626 Acquired: 6/25/2016 18:46:49 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00327	.00679	.00038	.03972	.02631	.00014	117.32	.00023
Stddev	.00072	.00538	.00107	.00258	.00068	.00005	.34	.00019
%RSD	22.022	79.326	284.61	6.5079	2.6006	35.377	.28683	83.944
#1	.00409	.00166	-.00067	.03695	.02674	.00014	117.70	.00003
#2	.00295	.01240	.00033	.04014	.02552	.00019	117.19	.00023
#3	.00276	.00630	.00147	.04206	.02666	.00009	117.07	.00042
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00020	.00126	.00604	.03025	1.5600	.01598	35.418	-.00162
Stddev	.00014	.00104	.00157	.00896	.0430	.00117	.155	.00177
%RSD	71.567	82.474	25.938	29.612	2.7589	7.3145	.43775	108.92
#1	.00004	.00238	.00775	.03493	1.5696	.01501	35.426	-.00126
#2	.00029	.00033	.00573	.01992	1.5974	.01567	35.569	-.00006
#3	.00026	.00107	.00466	.03588	1.5130	.01728	35.259	-.00354
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00045	17.994	-.00024	.00950	.00273	-.00253	.00429	2.8578
Stddev	.00029	.066	.00113	.00146	.00267	.00313	.00183	.0046
%RSD	65.363	.36928	461.23	15.318	98.010	123.57	42.598	.16060
#1	.00011	18.060	.00031	.00854	.00072	.00062	.00280	2.8530
#2	.00057	17.927	.00050	.00879	.00170	-.00564	.00374	2.8621
#3	.00066	17.996	-.00154	.01118	.00576	-.00258	.00633	2.8583
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098626 Acquired: 6/25/2016 18:46:49 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00012	.95352	-.01676	-.00181	.00025	.00656	.00282
Stddev	.00004	.00361	.00400	.00055	.00086	.00005	.00124
%RSD	36.480	.37812	23.874	30.252	341.08	.79230	44.156
#1	.00017	.95765	-.01632	-.00120	.00116	.00654	.00348
#2	.00013	.95102	-.02097	-.00198	-.00055	.00661	.00138
#3	.00008	.95189	-.01300	-.00225	.00014	.00651	.00358
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9709.6	59301.	3058.5
Stddev	45.7	137.	18.4
%RSD	.47038	.23115	.60138
#1	9758.2	59149.	3039.9
#2	9703.1	59336.	3058.7
#3	9667.6	59417.	3076.7

Approved: June 27, 2016

Sample Name: L1606098628 Acquired: 6/25/2016 18:50:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00127**	**.00062**	**.00226**	**.03408**	**.02631**	**.00011**	**202.60**
Stddev	.00142	.00505	.00320	.00146	.00145	.00006	1.09
%RSD	111.97	810.26	141.74	4.2925	5.4989	52.470	.53642
#1	.00192	-.00172	.00570	.03364	.02795	.00015	202.20
#2	.00225	-.00283	-.00063	.03289	.02521	.00012	201.76
#3	-.00036	.00642	.00171	.03571	.02576	.00005	203.83
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**-.00018**	**.00230**	**.00706**	**.02884**	**3.1108**	**.02231**
Stddev	.00015	.00008	.00109	.00066	.02764	.1253	.00468
%RSD	173.92	43.590	47.288	9.3696	95.862	4.0286	20.962
#1	.00015	-.00019	.00355	.00741	.06027	3.2199	.01700
#2	-.00008	-.00010	.00186	.00630	.00830	2.9739	.02412
#3	.00019	-.00026	.00151	.00747	.01794	3.1385	.02581
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**77.982**	**.03186**	**.00054**	**22.893**	**.00124**	**.00646**	**.00417**
Stddev	.421	.00093	.00029	.135	.00089	.00201	.00077
%RSD	.53979	2.9039	53.423	.59091	71.804	31.048	18.563
#1	77.763	.03137	.00064	22.869	.00109	.00439	.00500
#2	77.715	.03128	.00078	22.771	.00220	.00660	.00405
#3	78.467	.03292	.00022	23.039	.00044	.00839	.00347
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: L1606098628 Acquired: 6/25/2016 18:50:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00202**	**.00107**	**3.6233**	**.00060**	**2.4441**	F **-.03051**	**-.00323**
Stddev	.00386	.00287	.0015	.00175	.0120	.00406	.00089
%RSD	190.96	268.98	.04080	291.82	.49257	13.319	27.545
#1	.00200	.00365	3.6232	-.00046	2.4407	-.03513	-.00322
#2	-.00236	-.00203	3.6248	-.00036	2.4341	-.02890	-.00413
#3	-.00571	.00158	3.6218	.00262	2.4575	-.02750	-.00235
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00057**	**.01373**	**.00279**
Stddev	.00055	.00005	.00121
%RSD	96.088	.39414	43.503
#1	-.00086	.01376	.00139
#2	-.00092	.01376	.00355
#3	.00006	.01367	.00343
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9549.4**	**58061.**	**2994.7**
Stddev	10.8	342.	16.3
%RSD	.11322	.58988	.54589
#1	9556.3	57786.	2984.2
#2	9554.9	57953.	3013.5
#3	9536.9	58445.	2986.3

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 18:54:37 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.36142	9.0966	F .35164	F .43504	F .88841	.04556	F 8.8985
Stddev	.00063	.0241	.00214	.00359	.00668	.00015	.0636
%RSD	.17458	.26527	.60772	.82605	.75211	.32062	.71530
#1	.36081	9.1069	.35253	.43880	.88074	.04567	8.8250
#2	.36138	9.0691	.35318	.43467	.89150	.04539	8.9365
#3	.36207	9.1139	.34920	.43164	.89299	.04562	8.9341
Check ?	Chk Pass	Chk Pass	Chk Fail	Chk Fail	Chk Fail	Chk Pass	Chk Fail
Value			.40000	.50000	1.0000		10.000
Range			-10.000%	-10.000%	-10.000%		-10.000%

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04504	F .17683	F .44843	.45584	F 3.5682	F 43.223	.90012
Stddev	.00036	.00059	.00210	.00028	.0180	.402	.00642
%RSD	.79578	.33417	.46848	.06078	.50416	.93081	.71346
#1	.04493	.17615	.45006	.45571	3.5475	42.758	.89283
#2	.04475	.17711	.44916	.45565	3.5796	43.446	.90262
#3	.04544	.17723	.44606	.45616	3.5775	43.464	.90492
Check ?	Chk Pass	Chk Fail	Chk Fail	Chk Pass	Chk Fail	Chk Fail	Chk Pass
Value		.20000	.50000		4.0000	50.000	
Range		-10.000%	-10.000%		-10.000%	-10.000%	

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F 8.4568	F .44683	.90610	F 43.837	.45006	F 8.8106	F .44997
Stddev	.1941	.00153	.00307	.290	.00031	.0184	.00285
%RSD	2.2957	.34349	.33891	.66256	.06941	.20845	.63417
#1	8.2816	.44509	.90942	43.504	.44975	8.7897	.44790
#2	8.6655	.44738	.90552	43.973	.45037	8.8181	.44879
#3	8.4231	.44801	.90336	44.035	.45007	8.8241	.45322
Check ?	Chk Fail	Chk Fail	Chk Pass	Chk Fail	Chk Pass	Chk Fail	Chk Fail
Value	10.000	.50000		50.000		10.000	.50000
Range	-10.000%	-10.000%		-10.000%		-10.000%	-10.000%

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 18:54:37 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F 1.0511	F .35868	4.5430	F .88663	F .88982	F .89732	.46317
Stddev	.0038	.00393	.0064	.00114	.00700	.00126	.00115
%RSD	.35733	1.0963	.14052	.12909	.78679	.14039	.24883
#1	1.0514	.35522	4.5358	.88537	.88186	.89705	.46442
#2	1.0547	.35787	4.5453	.88691	.89256	.89622	.46295
#3	1.0472	.36295	4.5480	.88760	.89503	.89870	.46215
Check ?	Chk Fail	Chk Fail	Chk Pass	Chk Fail	Chk Fail	Chk Fail	Chk Pass
Value	1.2000	.40000		1.0000	1.0000	1.0000	
Range	-10.000%	-10.000%		-10.000%	-10.000%	-10.000%	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.92451	F .89326	F 2.6142
Stddev	.00239	.00079	.0303
%RSD	.25806	.08826	1.1591
#1	.92697	.89416	2.5794
#2	.92435	.89291	2.6285
#3	.92221	.89270	2.6347
Check ?	Chk Pass	Chk Fail	Chk Fail
Value		1.0000	1.0000
Range		-10.000%	10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10436.	62603.	3229.6
Stddev	7.	58.	39.3
%RSD	.06851	.09259	1.2179
#1	10435.	62597.	3274.8
#2	10443.	62548.	3203.4
#3	10429.	62663.	3210.5

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 18:58:15 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00102	-.00634	.00315	-.00143	.00123	.00009	-.03120	-.00001
Stddev	.00036	.00615	.00146	.00223	.00014	.00007	.03757	.00022
%RSD	35.623	96.910	46.271	156.16	11.328	79.889	120.40	3000.5
#1	.00142	-.00623	.00470	-.00112	.00129	.00016	.01024	.00017
#2	.00091	-.01255	.00293	-.00379	.00133	.00004	-.06303	-.00025
#3	.00072	-.00026	.00181	.00063	.00107	.00005	-.04081	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00001	-.00012	.00040	.00652	.00729	-.00094	.01960	-.00009
Stddev	.00003	.00070	.00057	.00448	.10732	.00258	.12813	.00039
%RSD	340.66	581.54	142.03	68.614	1471.3	275.05	653.72	408.22
#1	-.00000	.00019	-.00025	.00247	.06763	-.00204	-.01079	.00011
#2	.00004	-.00093	.00068	.01133	.07087	.00201	.16019	-.00054
#3	-.00001	.00037	.00077	.00577	-.11662	-.00279	-.09060	.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00312	.01765	-.00021	.00497	.00440	.00279	.00542	.03094
Stddev	.00023	.01443	.00123	.00520	.00590	.00458	.00181	.00198
%RSD	7.4573	81.775	586.87	104.64	134.12	164.15	33.357	6.4039
#1	.00298	.03423	-.00101	.00374	.00273	.00293	.00481	.03016
#2	.00339	.00793	-.00083	.01067	.01095	-.00186	.00399	.03319
#3	.00300	.01078	.00121	.00049	-.00049	.00730	.00745	.02947
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 18:58:15 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00029**	**.00057**	**.00328**	**-.00363**	**.00052**	**.00010**	**.00222**
Stddev	.00020	.00050	.00575	.00205	.00106	.00011	.00064
%RSD	66.832	88.088	175.09	56.396	206.25	116.18	28.891
#1	.00017	.00114	-.00137	-.00544	-.00047	.00019	.00164
#2	.00052	.00028	.00971	-.00141	.00038	-.00003	.00291
#3	.00019	.00027	.00151	-.00405	.00164	.00012	.00210
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10029.**	**60890.**	**3064.8**
Stddev	14.	97.	13.9
%RSD	.13680	.15850	.45397
#1	10045.	60780.	3055.3
#2	10022.	60930.	3080.8
#3	10020.	60960.	3058.3

Approved: June 27, 2016

Sample Name: L1606098630 Acquired: 6/25/2016 19:02:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00160**	**.04734**	**-.00156**	**.13569**	**.01871**	**.00015**	**238.04**
Stddev	.00181	.00348	.00404	.00135	.00123	.00008	.85
%RSD	113.48	7.3553	258.52	.99665	6.5967	52.824	.35679
#1	.00017	.04905	-.00155	.13421	.01822	.00022	239.01
#2	.00364	.04964	.00247	.13685	.01779	.00007	237.71
#3	.00098	.04333	-.00561	.13603	.02011	.00014	237.41
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00026**	**.00032**	**.00251**	**.01182**	**.03060**	**3.4183**	**.04090**
Stddev	.00010	.00018	.00081	.00072	.01332	.0462	.00200
%RSD	39.577	56.009	32.204	6.0677	43.540	1.3506	4.8796
#1	.00014	.00047	.00287	.01260	.01963	3.3856	.03874
#2	.00034	.00036	.00308	.01119	.04543	3.4711	.04128
#3	.00028	.00012	.00159	.01167	.02674	3.3981	.04268
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**90.439**	**.21655**	**.00156**	**78.879**	**.00140**	**.01756**	**.00033**
Stddev	.173	.00121	.00013	.416	.00068	.00677	.00046
%RSD	.19086	.55720	8.0204	.52709	48.349	38.536	138.30
#1	90.626	.21726	.00143	79.353	.00191	.01647	.00076
#2	90.405	.21515	.00168	78.710	.00063	.02480	-.00016
#3	90.286	.21722	.00157	78.574	.00166	.01140	.00040
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: L1606098630 Acquired: 6/25/2016 19:02:11 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00294**	**-.00177**	**5.6111**	**-.00017**	**7.0275**	F **-.03316**	**-.00254**
Stddev	.00229	.00087	.0076	.00038	.0379	.00293	.00099
%RSD	77.743	49.294	.13521	227.38	.53946	8.8247	38.798
#1	-.00354	-.00250	5.6121	-.00026	7.0707	-.03043	-.00171
#2	-.00042	-.00199	5.6030	-.00050	7.0124	-.03625	-.00363
#3	-.00488	-.00081	5.6181	.00025	6.9995	-.03280	-.00229
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00065**	**.02003**	**.00667**
Stddev	.00064	.00007	.00116
%RSD	98.374	.34297	17.407
#1	-.00116	.01996	.00557
#2	-.00088	.02003	.00654
#3	.00007	.02010	.00788
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9361.9**	**56594.**	**2990.6**
Stddev	23.4	147.	20.1
%RSD	.24995	.25968	.67365
#1	9379.7	56762.	2967.6
#2	9370.7	56528.	3005.2
#3	9335.4	56491.	2998.9

Approved: June 27, 2016

Sample Name: L1606098632S Acquired: 6/25/2016 19:06:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-06

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.18902**	**4.8212**	**.19116**	**.97374**	**.50077**	**.02402**	**134.52**	**.02388**
Stddev	.00088	.0081	.00239	.00462	.00118	.00003	.34	.00017
%RSD	.46488	.16867	1.2481	.47438	.23617	.14314	.25228	.70472
#1	.18980	4.8271	.19377	.97198	.50123	.02406	134.68	.02373
#2	.18919	4.8246	.18910	.97898	.49943	.02400	134.13	.02406
#3	.18807	4.8119	.19060	.97026	.50165	.02400	134.75	.02386
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09189**	**.23665**	**.24218**	**1.8859**	**25.617**	**.51395**	**37.596**	**.28166**
Stddev	.00021	.00120	.00028	.0089	.037	.00043	.192	.00203
%RSD	.23170	.50673	.11712	.47202	.14467	.08375	.50984	.71945
#1	.09190	.23534	.24190	1.8828	25.580	.51444	37.585	.28070
#2	.09167	.23691	.24247	1.8960	25.619	.51376	37.792	.28399
#3	.09210	.23769	.24219	1.8790	25.654	.51365	37.409	.28029
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49448**	**71.230**	**.23175**	**4.7310**	**.23416**	**.56777**	**.18400**	**6.1097**
Stddev	.00088	.164	.00128	.0060	.00191	.00602	.00319	.0038
%RSD	.17884	.22978	.55101	.12688	.81554	1.0604	1.7354	.06148
#1	.49529	71.252	.23067	4.7327	.23616	.57278	.18033	6.1129
#2	.49354	71.056	.23316	4.7244	.23394	.56944	.18613	6.1108
#3	.49461	71.382	.23142	4.7360	.23236	.56109	.18555	6.1056
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098632S Acquired: 6/25/2016 19:06:04 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-06

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.46869	2.2839	.45630	.23531	.49366	.46774	.01477
Stddev	.00224	.0055	.00571	.00163	.00079	.00046	.00310
%RSD	.47718	.23955	1.2515	.69229	.16011	.09871	21.022
#1	.47047	2.2847	.45118	.23550	.49390	.46819	.01151
#2	.46942	2.2781	.45528	.23360	.49431	.46777	.01510
#3	.46618	2.2889	.46246	.23684	.49278	.46726	.01769
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9505.1	57443.	2995.1
Stddev	12.6	181.	7.9
%RSD	.13263	.31567	.26290
#1	9490.9	57238.	2989.7
#2	9509.8	57583.	3004.2
#3	9514.8	57507.	2991.6

Approved: June 27, 2016

Sample Name: L1606098634S Acquired: 6/25/2016 19:09:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19165**	**4.8639**	**.19259**	**.98224**	**.50433**	**.02434**	**136.41**	**.02397**
Stddev	.00283	.0058	.00215	.00673	.00181	.00002	.41	.00039
%RSD	1.4742	.11932	1.1143	.68564	.35885	.08391	.29851	1.6453
#1	.19474	4.8576	.19410	.97512	.50399	.02434	136.06	.02438
#2	.19099	4.8690	.19014	.98851	.50272	.02431	136.30	.02394
#3	.18921	4.8651	.19354	.98310	.50629	.02435	136.86	.02359
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09370**	**.23944**	**.24457**	**1.9014**	**25.851**	**.51838**	**38.336**	**.27292**
Stddev	.00027	.00087	.00143	.0128	.044	.00156	.167	.00356
%RSD	.28910	.36336	.58497	.67456	.17178	.30011	.43499	1.3028
#1	.09395	.23880	.24603	1.9045	25.902	.51991	38.172	.27676
#2	.09341	.23909	.24449	1.9124	25.822	.51842	38.330	.27226
#3	.09373	.24043	.24318	1.8873	25.829	.51680	38.505	.26974
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49947**	**74.848**	**.23412**	**4.7760**	**.23373**	**.57493**	**.18272**	**6.1711**
Stddev	.00101	.224	.00114	.0149	.00097	.00427	.00635	.0140
%RSD	.20163	.29965	.48754	.31111	.41349	.74220	3.4752	.22725
#1	.50053	74.661	.23450	4.7820	.23289	.57819	.18571	6.1610
#2	.49852	74.786	.23283	4.7591	.23352	.57010	.18702	6.1652
#3	.49936	75.096	.23502	4.7869	.23478	.57650	.17542	6.1871
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098634S Acquired: 6/25/2016 19:09:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-07

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47482**	**2.3307**	**.45927**	**.23900**	**.49848**	**.47191**	**.00533**
Stddev	.00086	.0091	.01368	.00122	.00164	.00108	.00076
%RSD	.18068	.39132	2.9792	.51192	.32871	.22881	14.237
#1	.47581	2.3242	.47279	.23871	.49842	.47284	.00492
#2	.47440	2.3268	.44543	.23794	.50015	.47073	.00620
#3	.47425	2.3411	.45961	.24034	.49687	.47217	.00485
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9464.2**	**57182.**	**2975.2**
Stddev	11.1	55.	14.2
%RSD	.11755	.09672	.47580
#1	9463.1	57128.	2991.6
#2	9475.9	57180.	2967.1
#3	9453.7	57239.	2967.1

Approved: June 27, 2016

Sample Name: L1606098636SD Acquired: 6/25/2016 19:13:21 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19162**	**4.8756**	**.19532**	**1.3252**	**.50705**	**.02431**	**243.19**	**.02436**
Stddev	.00178	.0135	.00259	.0039	.00540	.00016	2.22	.00014
%RSD	.92905	.27765	1.3247	.29569	1.0646	.65896	.91281	.56637
#1	.19240	4.8805	.19500	1.3207	.51326	.02435	245.71	.02449
#2	.18958	4.8859	.19805	1.3282	.50441	.02444	242.33	.02421
#3	.19287	4.8603	.19291	1.3265	.50349	.02413	241.52	.02439
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09290**	**.23671**	**.29685**	**1.9145**	**29.596**	**.51587**	**71.839**	**.60737**
Stddev	.00022	.00172	.00059	.0369	.191	.00194	.252	.00417
%RSD	.23161	.72762	.19867	1.9280	.64372	.37694	.35080	.68663
#1	.09280	.23499	.29694	1.9564	29.779	.51781	72.039	.61218
#2	.09315	.23843	.29739	1.9005	29.610	.51588	71.921	.60495
#3	.09276	.23672	.29622	1.8867	29.399	.51393	71.556	.60497
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49958**	**89.482**	**.23320**	**4.8335**	**.23655**	**.57885**	**.18210**	**6.3388**
Stddev	.00153	.784	.00050	.0076	.00169	.00511	.00473	.0094
%RSD	.30585	.87569	.21296	.15803	.71424	.88252	2.5997	.14817
#1	.49951	90.366	.23335	4.8331	.23580	.57825	.18575	6.3355
#2	.50114	89.209	.23361	4.8414	.23849	.57407	.18379	6.3494
#3	.49809	88.872	.23265	4.8261	.23537	.58424	.17675	6.3314
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098636SD Acquired: 6/25/2016 19:13:21 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-08

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47318**	**2.1504**	**.45196**	**.23380**	**.49243**	**.47726**	**.00396**
Stddev	.00235	.0203	.01296	.00229	.00254	.00239	.00140
%RSD	.49768	.94209	2.8683	.97909	.51630	.50171	35.414
#1	.47220	2.1736	.45301	.23117	.49302	.48002	.00408
#2	.47587	2.1417	.46436	.23534	.49462	.47583	.00530
#3	.47148	2.1359	.43850	.23489	.48964	.47592	.00250
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9466.5**	**57687.**	**3000.7**
Stddev	28.5	42.	21.4
%RSD	.30105	.07212	.71153
#1	9499.3	57706.	2976.5
#2	9448.1	57639.	3017.0
#3	9452.0	57716.	3008.4

Approved: June 27, 2016

Sample Name: L1606098638SD Acquired: 6/25/2016 19:16:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-09

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19635**	**5.0111**	**.19662**	**1.3658**	**.50590**	**.02455**	**238.74**	**.02470**
Stddev	.00189	.0495	.00202	.0129	.00507	.00013	2.44	.00046
%RSD	.96164	.98714	1.0288	.94297	1.0027	.51871	1.0237	1.8632
#1	.19421	4.9540	.19875	1.3513	.50004	.02440	236.05	.02465
#2	.19777	5.0396	.19638	1.3761	.50894	.02464	240.82	.02518
#3	.19706	5.0398	.19472	1.3699	.50872	.02461	239.35	.02426
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09456**	**.24315**	**.24888**	**1.9280**	**29.506**	**.51676**	**69.466**	**.60260**
Stddev	.00054	.00286	.00145	.0328	.362	.00473	.757	.00806
%RSD	.56749	1.1744	.58163	1.6990	1.2271	.91552	1.0898	1.3377
#1	.09518	.24020	.24797	1.8914	29.119	.51186	68.628	.59529
#2	.09432	.24590	.25055	1.9547	29.837	.52131	70.101	.61124
#3	.09419	.24336	.24812	1.9378	29.562	.51712	69.668	.60126
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50828**	**86.839**	**.23731**	**4.9079**	**.23796**	**.58722**	**.18727**	**6.3571**
Stddev	.00153	.812	.00040	.0105	.00122	.00763	.00717	.0147
%RSD	.30188	.93521	.16702	.21476	.51147	1.2987	3.8265	.23138
#1	.50998	85.938	.23758	4.9185	.23889	.59365	.18466	6.3682
#2	.50784	87.516	.23685	4.9076	.23841	.58922	.19537	6.3627
#3	.50701	87.061	.23748	4.8975	.23659	.57879	.18177	6.3405
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098638SD Acquired: 6/25/2016 19:16:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573530-09

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48230**	**2.1413**	**.45206**	**.24148**	**.50819**	**.47197**	**.00465**
Stddev	.00016	.0221	.00841	.00203	.00480	.00162	.00175
%RSD	.03337	1.0323	1.8614	.84256	.94539	.34380	37.692
#1	.48228	2.1164	.44367	.23931	.50264	.47335	.00633
#2	.48215	2.1586	.45201	.24177	.51108	.47237	.00283
#3	.48247	2.1488	.46050	.24334	.51084	.47018	.00481
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9367.1**	**56617.**	**3005.4**
Stddev	24.2	279.	8.9
%RSD	.25886	.49353	.29743
#1	9345.0	56900.	3015.6
#2	9363.2	56609.	2998.7
#3	9393.0	56341.	3002.0

Approved: June 27, 2016

Sample Name: L1606098640 Acquired: 6/25/2016 19:20:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00070	.00014	.00047	.01204	.00394	.00013	2.2418	.00009
Stddev	.00052	.00246	.00354	.00206	.00041	.00015	.0448	.00026
%RSD	75.327	1722.8	753.06	17.114	10.437	111.96	1.9977	298.05
#1	-.00120	.00298	.00434	.01328	.00360	.00001	2.2906	-.00017
#2	-.00073	-.00107	-.00032	.00966	.00384	.00009	2.2025	.00008
#3	-.00016	-.00148	-.00261	.01318	.00440	.00029	2.2324	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00011	-.00023	.00162	.03197	.19488	-.00199	.70657	-.00157
Stddev	.00020	.00043	.00063	.03003	.05117	.00159	.10487	.00034
%RSD	172.28	188.43	38.822	93.912	26.255	79.698	14.843	21.723
#1	-.00014	.00024	.00129	.06510	.20250	-.00219	.77648	-.00118
#2	.00010	-.00061	.00122	.00654	.14033	-.00346	.58598	-.00170
#3	-.00030	-.00032	.00234	.02429	.24180	-.00031	.75725	-.00182
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00061	1.3245	-.00000	-.00132	.00155	-.00034	.00103	1.8661
Stddev	.00043	.0047	.00128	.00539	.00283	.00334	.00522	.0030
%RSD	69.622	.35334	42382.	408.47	183.18	979.63	508.03	.16155
#1	.00110	1.3264	-.00135	.00378	-.00057	-.00417	-.00062	1.8696
#2	.00029	1.3279	.00121	-.00695	.00477	.00193	.00687	1.8648
#3	.00045	1.3192	.00013	-.00078	.00045	.00122	-.00317	1.8641
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098640 Acquired: 6/25/2016 19:20:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00060	.02890	-.00018	-.00245	.00125	.00316	.00092
Stddev	.00104	.00022	.00313	.00122	.00036	.00008	.00187
%RSD	172.87	.75581	1755.3	49.848	28.928	2.5446	202.44
#1	-.00053	.02865	.00335	-.00105	.00165	.00322	-.00102
#2	.00152	.02900	-.00128	-.00328	.00095	.00320	.00271
#3	.00082	.02905	-.00260	-.00302	.00115	.00307	.00108
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9902.6	60698.	3047.4
Stddev	11.3	88.	21.1
%RSD	.11376	.14441	.69149
#1	9915.6	60659.	3066.8
#2	9895.9	60798.	3025.0
#3	9896.3	60636.	3050.3

Approved: June 27, 2016

Sample Name: L1606098642 Acquired: 6/25/2016 19:24:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00084	-.00125	.00012	.00714	.00291	.00014	2.2957	-.00008
Stddev	.00151	.00416	.00315	.00149	.00136	.00010	.0287	.00011
%RSD	179.62	331.72	2535.0	20.853	46.790	76.580	1.2503	140.66
#1	.00086	.00295	-.00350	.00543	.00443	.00023	2.2645	-.00020
#2	-.00134	-.00134	.00221	.00806	.00180	.00015	2.3210	.00001
#3	-.00204	-.00537	.00167	.00794	.00250	.00002	2.3016	-.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00027	.00054	.00126	.01879	.13459	.00101	.70251	-.00031
Stddev	.00021	.00073	.00033	.01779	.06299	.00139	.04015	.00188
%RSD	76.216	135.58	26.342	94.669	46.805	137.70	5.7150	613.21
#1	.00043	.00109	.00162	.03687	.13442	.00244	.65916	.00151
#2	.00035	-.00029	.00096	.01820	.19766	.00092	.73842	-.00224
#3	.00004	.00081	.00121	.00130	.07167	-.00034	.70994	-.00020
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00012	1.3914	-.00012	.00599	-.00050	-.00197	.00192	1.9147
Stddev	.00031	.0274	.00128	.00446	.00248	.00478	.00405	.0013
%RSD	265.01	1.9704	1039.6	74.477	491.66	243.04	211.10	.06740
#1	-.00017	1.4067	.00078	.00869	.00158	.00334	.00542	1.9141
#2	.00008	1.4078	.00044	.00843	.00015	-.00331	-.00251	1.9137
#3	.00044	1.3598	-.00159	.00084	-.00325	-.00593	.00284	1.9161
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098642 Acquired: 6/25/2016 19:24:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**.02978**	**.00659**	**-.00224**	**.00021**	**.00288**	**.00046**
Stddev	.00041	.00031	.00898	.00218	.00024	.00011	.00203
%RSD	319.77	1.0309	136.22	97.103	115.48	3.6974	442.65
#1	.00034	.02969	.01094	-.00308	.00030	.00286	-.00018
#2	-.00043	.03012	.01258	.00023	.00040	.00300	.00273
#3	-.00030	.02953	-.00373	-.00388	-.00006	.00279	-.00117
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9910.0**	**60405.**	**3027.1**
Stddev	8.6	109.	25.1
%RSD	.08719	.18086	.83024
#1	9909.0	60400.	3001.0
#2	9919.1	60298.	3029.3
#3	9901.9	60516.	3051.1

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 19:28:27 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.38218	9.6308	.37557	.46431	.95962	.04828	9.5695
Stddev	.00218	.0178	.00127	.00116	.00217	.00012	.0717
%RSD	.57165	.18471	.33835	.24921	.22635	.23895	.74942
#1	.37992	9.6388	.37606	.46299	.96032	.04841	9.5557
#2	.38232	9.6104	.37652	.46516	.96137	.04820	9.6471
#3	.38429	9.6432	.37413	.46479	.95719	.04823	9.5057
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.04801	.18801	.47395	.48473	3.7689	46.355	.97311
Stddev	.00014	.00020	.00045	.00153	.0264	.220	.00378
%RSD	.29427	.10445	.09483	.31643	.70136	.47512	.38849
#1	.04785	.18823	.47438	.48634	3.7952	46.500	.97746
#2	.04812	.18785	.47399	.48328	3.7693	46.464	.97068
#3	.04806	.18795	.47348	.48458	3.7423	46.102	.97118
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.0717	.47590	.96524	47.107	.47852	9.3733	.47882
Stddev	.0608	.00396	.00399	.173	.00072	.0111	.00296
%RSD	.67059	.83199	.41378	.36832	.15104	.11884	.61810
#1	9.0038	.47335	.96959	47.188	.47773	9.3765	.47551
#2	9.1213	.48046	.96440	47.226	.47916	9.3609	.48123
#3	9.0900	.47388	.96173	46.908	.47866	9.3825	.47971
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 19:28:27 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1223**	**.37795**	**4.8400**	**.94463**	**.95969**	**.96684**	**.48966**
Stddev	.0016	.00541	.0100	.00151	.00470	.01176	.00307
%RSD	.14404	1.4313	.20631	.15941	.49016	1.2167	.62598
#1	1.1238	.37410	4.8492	.94636	.96229	.96908	.48613
#2	1.1206	.37562	4.8414	.94360	.96253	.97732	.49169
#3	1.1224	.38414	4.8294	.94394	.95426	.95412	.49116
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.97958**	**.95235**	F **2.8246**
Stddev	.00154	.00042	.0035
%RSD	.15724	.04377	.12384
#1	.97882	.95263	2.8278
#2	.97857	.95254	2.8252
#3	.98136	.95187	2.8208
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10132.**	**61264.**	**3100.4**
Stddev	.	134.	6.9
%RSD	.00202	.21941	.22117
#1	10132.	61202.	3092.6
#2	10132.	61418.	3103.3
#3	10132.	61171.	3105.3

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 19:32:05 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00027**	**-.00034**	**.00165**	**-.00065**	**.00143**	**.00014**	**-.04509**	**-.00015**
Stddev	.00053	.00835	.00116	.00135	.00075	.00012	.04632	.00018
%RSD	198.69	2437.1	70.395	208.11	52.425	84.191	102.74	120.05
#1	.00081	.00499	.00229	-.00108	.00058	.00001	-.09453	-.00022
#2	.00023	-.00997	.00235	-.00174	.00173	.00020	-.03804	.00005
#3	-.00025	.00395	.00031	.00087	.00199	.00023	-.00270	-.00028
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00008**	**-.00017**	**.00065**	**.02423**	**-.06433**	**-.00295**	**-.01903**	**-.00066**
Stddev	.00013	.00077	.00059	.01555	.07126	.00237	.08401	.00260
%RSD	162.10	456.13	90.441	64.178	110.77	80.373	441.41	395.33
#1	-.00022	-.00087	-.00002	.03897	.01654	-.00026	-.10302	.00036
#2	-.00006	.00066	.00091	.02575	-.09166	-.00387	.06499	-.00361
#3	.00004	-.00030	.00106	.00798	-.11789	-.00472	-.01907	.00127
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00313**	**.02145**	**.00120**	**.00197**	**.00164**	**.00212**	**.00101**	**.04112**
Stddev	.00013	.00743	.00050	.00366	.00312	.00166	.00069	.00272
%RSD	4.1237	34.660	41.579	185.66	190.04	78.472	68.223	6.6255
#1	.00318	.03002	.00170	.00620	.00310	.00128	.00172	.03946
#2	.00323	.01754	.00070	-.00004	.00377	.00104	.00035	.04426
#3	.00298	.01678	.00121	-.00024	-.00194	.00404	.00097	.03964
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 19:32:05 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00072**	**.00051**	**-.00006**	**-.00149**	**-.00026**	**.00006**	**.00316**
Stddev	.00004	.00041	.00466	.00123	.00016	.00032	.00144
%RSD	4.9365	80.523	8195.1	82.501	61.089	521.42	45.405
#1	.00073	.00010	.00498	-.00182	-.00044	-.00016	.00318
#2	.00068	.00091	-.00094	-.00252	-.00020	.00042	.00172
#3	.00075	.00050	-.00421	-.00013	-.00014	-.00008	.00459
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9625.9**	**58446.**	**2947.7**
Stddev	17.4	102.	8.2
%RSD	.18036	.17465	.27941
#1	9605.9	58491.	2954.6
#2	9635.9	58518.	2950.0
#3	9636.0	58329.	2938.6

Approved: June 27, 2016

Sample Name: PBW LH Acquired: 6/25/2016 19:36:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00022	-.00480	.00297	-.00088	.00071	.00016	-.03609	-.00030
Stddev	.00096	.00538	.00061	.00047	.00081	.00006	.00576	.00022
%RSD	436.54	112.05	20.402	53.250	112.94	41.040	15.975	74.291
#1	-.00063	-.00271	.00258	-.00071	.00057	.00010	-.04014	-.00038
#2	.00002	-.00078	.00267	-.00141	-.00001	.00022	-.03863	-.00005
#3	.00126	-.01092	.00367	-.00052	.00158	.00014	-.02949	-.00047
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00009	.00027	.00021	.03551	.02057	.00058	-.03422	.00199
Stddev	.00019	.00046	.00040	.00062	.11033	.00127	.05831	.00213
%RSD	211.66	173.58	192.44	1.7348	536.36	218.26	170.39	107.33
#1	.00009	.00043	-.00014	.03559	.14793	.00205	.01703	.00051
#2	-.00029	.00063	.00065	.03485	-.04548	-.00019	-.09766	.00101
#3	-.00007	-.00025	.00012	.03607	-.04075	-.00011	-.02203	.00443
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00027	.01920	-.00013	.00015	.00308	-.00169	.00218	.00530
Stddev	.00020	.01381	.00043	.00436	.00410	.00157	.00344	.00135
%RSD	72.649	71.919	323.98	2952.6	133.25	92.573	157.59	25.417
#1	-.00050	.00540	-.00047	-.00185	.00268	-.00008	.00514	.00605
#2	-.00014	.01919	.00035	-.00286	.00737	-.00321	.00301	.00374
#3	-.00017	.03301	-.00027	.00515	-.00081	-.00179	-.00160	.00610
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: PBW LH Acquired: 6/25/2016 19:36:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00038**	**.00001**	**.00007**	**-.00205**	**.00056**	**-.00012**	**.00138**
Stddev	.00105	.00006	.00526	.00274	.00041	.00008	.00122
%RSD	278.35	503.89	7708.4	133.89	73.183	68.172	88.458
#1	.00057	-.00005	.00045	-.00401	.00011	-.00018	.00093
#2	-.00020	.00002	-.00537	-.00321	.00065	-.00003	.00045
#3	-.00151	.00007	.00513	.00108	.00091	-.00016	.00276
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10355.**	**62690.**	**3143.1**
Stddev	14.	21.	40.1
%RSD	.13810	.03356	1.2743
#1	10371.	62711.	3176.8
#2	10343.	62669.	3098.8
#3	10351.	62689.	3153.6

Approved: June 27, 2016

Sample Name: LCSW LH Acquired: 6/25/2016 19:39:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19366**	**4.9137**	**.19220**	**.93203**	**.49068**	**.02467**	**4.9632**	**.02443**
Stddev	.00070	.0138	.00341	.00396	.00423	.00007	.0538	.00016
%RSD	.36335	.28142	1.7740	.42472	.86189	.28277	1.0835	.64799
#1	.19288	4.8983	.19002	.92766	.48616	.02461	4.9413	.02459
#2	.19385	4.9180	.19044	.93305	.49453	.02465	5.0244	.02441
#3	.19425	4.9249	.19613	.93537	.49136	.02474	4.9238	.02428
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09715**	**.24544**	**.24999**	**1.9443**	**23.472**	**.50293**	**4.6813**	**.24132**
Stddev	.00066	.00263	.00041	.0106	.182	.00291	.0938	.00318
%RSD	.68088	1.0726	.16565	.54556	.77586	.57762	2.0038	1.3188
#1	.09680	.24301	.25039	1.9373	23.301	.50365	4.6704	.23802
#2	.09792	.24824	.24956	1.9391	23.664	.50540	4.7800	.24157
#3	.09674	.24506	.25002	1.9565	23.450	.49973	4.5934	.24437
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50304**	**24.148**	**.24571**	**4.6828**	**.24554**	**.57884**	**.18202**	**2.4064**
Stddev	.00049	.122	.00061	.0125	.00123	.00461	.00367	.0046
%RSD	.09743	.50324	.24625	.26789	.50266	.79649	2.0171	.19316
#1	.50339	24.028	.24640	4.6758	.24587	.57352	.17868	2.4078
#2	.50325	24.271	.24545	4.6973	.24658	.58167	.18595	2.4102
#3	.50248	24.145	.24528	4.6753	.24418	.58132	.18143	2.4012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LCSW LH Acquired: 6/25/2016 19:39:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48493**	**.49684**	**.48674**	**.24907**	**.50464**	**.48284**	**.01676**
Stddev	.00108	.00397	.00310	.00041	.00236	.00325	.00211
%RSD	.22287	.79928	.63744	.16637	.46686	.67256	12.564
#1	.48462	.49378	.48665	.24947	.50200	.48055	.01435
#2	.48613	.50133	.48369	.24909	.50538	.48656	.01775
#3	.48404	.49542	.48989	.24864	.50653	.48141	.01820
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9761.3**	**59171.**	**3026.9**
Stddev	5.3	58.	23.7
%RSD	.05400	.09751	.78195
#1	9766.4	59236.	3050.7
#2	9755.9	59148.	3003.3
#3	9761.7	59127.	3026.7

Approved: June 27, 2016

Sample Name: L1606090723 Acquired: 6/25/2016 19:43:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00091	.01800	.14333	.20159	.04322	.00011	82.135	.00062
Stddev	.00074	.00440	.00262	.00108	.00176	.00009	.153	.00002
%RSD	81.113	24.457	1.8284	.53531	4.0750	85.790	.18660	3.6806
#1	.00039	.02274	.14560	.20094	.04396	.00020	82.043	.00061
#2	.00058	.01723	.14046	.20283	.04121	.00011	82.051	.00064
#3	.00176	.01403	.14393	.20099	.04450	.00001	82.312	.00060
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00085	.00186	.00181	3.0759	4.3146	.03826	51.383	.03252
Stddev	.00009	.00076	.00070	.0448	.0921	.00398	.083	.00353
%RSD	10.346	41.113	38.683	1.4576	2.1337	10.401	.16137	10.870
#1	.00081	.00098	.00242	3.0253	4.2316	.04144	51.475	.03488
#2	.00094	.00223	.00197	3.0916	4.4136	.03380	51.362	.02845
#3	.00078	.00237	.00105	3.1108	4.2984	.03956	51.313	.03422
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00258	133.73	-.00015	.01191	.00179	-.00228	.00665	8.0891
Stddev	.00008	.12	.00113	.00694	.00129	.00283	.00532	.0280
%RSD	3.2153	.09189	747.26	58.268	71.933	124.12	79.994	.34670
#1	.00264	133.81	-.00126	.01916	.00318	-.00553	.01195	8.1130
#2	.00249	133.59	.00099	.00533	.00159	-.00033	.00671	8.0582
#3	.00262	133.80	-.00018	.01123	.00062	-.00099	.00131	8.0960
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606090723 Acquired: 6/25/2016 19:43:38 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00001**	**2.6416**	**-.00732**	**-.00182**	**-.00046**	**.00542**	**.00128**
Stddev	.00038	.0023	.00771	.00082	.00062	.00021	.00019
%RSD	7616.2	.08543	105.44	45.025	136.48	3.8805	14.578
#1	.00020	2.6400	-.01112	-.00260	-.00073	.00544	.00120
#2	.00025	2.6406	.00156	-.00189	.00026	.00563	.00115
#3	-.00044	2.6441	-.01239	-.00097	-.00089	.00521	.00149
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9566.3**	**57706.**	**3016.3**
Stddev	21.2	40.	8.6
%RSD	.22170	.06981	.28561
#1	9583.7	57737.	3006.9
#2	9572.5	57660.	3023.9
#3	9542.7	57721.	3018.2

Approved: June 27, 2016

Sample Name: L1606098601 Acquired: 6/25/2016 19:47:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00061	.63226	.00101	.09520	.04126	.00012	31.325	.00014
Stddev	.00183	.00734	.00119	.00196	.00050	.00007	.135	.00021
%RSD	298.65	1.1612	118.08	2.0630	1.2197	57.984	.43030	153.25
#1	.00185	.62395	-.00028	.09744	.04176	.00005	31.451	-.00001
#2	-.00149	.63786	.00207	.09436	.04076	.00018	31.183	.00005
#3	.00148	.63498	.00124	.09379	.04125	.00013	31.342	.00038
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00117	.00251	.00378	1.4621	1.2075	.01677	5.9028	.04892
Stddev	.00024	.00020	.00199	.0362	.0294	.00358	.0605	.00066
%RSD	20.175	7.8011	52.767	2.4785	2.4351	21.330	1.0243	1.3538
#1	.00129	.00246	.00338	1.4981	1.1821	.01735	5.9016	.04861
#2	.00090	.00272	.00594	1.4256	1.2008	.02003	5.8430	.04848
#3	.00133	.00234	.00201	1.4624	1.2397	.01294	5.9639	.04969
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00119	122.45	.00122	.03952	.00058	-.00189	.00677	4.0503
Stddev	.00024	.35	.00124	.00581	.00157	.00112	.00463	.0371
%RSD	19.925	.28250	101.52	14.694	270.52	59.209	68.294	.91471
#1	.00135	122.68	.00181	.04077	.00200	-.00256	.00933	4.0789
#2	.00131	122.05	-.00020	.04460	-.00110	-.00252	.00956	4.0634
#3	.00092	122.62	.00206	.03319	.00084	-.00060	.00143	4.0084
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098601 Acquired: 6/25/2016 19:47:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00016**	**.54428**	**.01566**	**-.00252**	**.00147**	**.03503**	**.00138**
Stddev	.00052	.00100	.00262	.00158	.00115	.00040	.00180
%RSD	330.58	.18426	16.757	62.766	78.497	1.1335	130.15
#1	-.00032	.54517	.01814	-.00107	.00112	.03533	.00331
#2	-.00057	.54319	.01592	-.00228	.00275	.03517	-.00024
#3	.00042	.54448	.01291	-.00422	.00053	.03458	.00107
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9674.1**	**58435.**	**3018.4**
Stddev	45.2	129.	18.2
%RSD	.46720	.22152	.60258
#1	9673.0	58340.	2997.4
#2	9719.9	58382.	3028.7
#3	9629.5	58582.	3029.1

Approved: June 27, 2016

Sample Name: L1606098603 Acquired: 6/25/2016 19:51:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00113**	**.08091**	**.00257**	**.02612**	**.02343**	**.00026**	**258.72**
Stddev	.00364	.00403	.00338	.00065	.00023	.00010	.22
%RSD	322.73	4.9793	131.60	2.5002	.99712	37.911	.08495
#1	-.00201	.07953	.00151	.02617	.02317	.00019	258.95
#2	.00512	.08545	-.00015	.02675	.02363	.00037	258.52
#3	.00028	.07776	.00635	.02545	.02348	.00021	258.68
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00056**	**.00486**	**.00101**	**.01760**	**1.6277**	**1.7887**	**.01926**
Stddev	.00018	.00049	.00101	.00048	.0214	.0748	.00269
%RSD	32.935	10.099	100.65	2.7241	1.3157	4.1799	13.963
#1	.00053	.00446	.00070	.01814	1.6519	1.7719	.02232
#2	.00040	.00472	.00018	.01725	1.6198	1.8705	.01729
#3	.00076	.00541	.00214	.01740	1.6113	1.7238	.01816
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**67.571**	**3.7591**	**.00091**	**16.176**	**.00450**	**.03792**	**.00200**
Stddev	.265	.0091	.00013	.042	.00050	.00459	.00102
%RSD	.39176	.24313	14.640	.25963	11.081	12.099	51.049
#1	67.875	3.7689	.00083	16.148	.00428	.03292	.00317
#2	67.391	3.7508	.00083	16.224	.00415	.03891	.00158
#3	67.448	3.7575	.00106	16.157	.00507	.04193	.00126
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: L1606098603 Acquired: 6/25/2016 19:51:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00201**	**-.00717**	**3.9929**	**.00071**	**1.4034**	F **-.03512**	**-.00302**
Stddev	.00033	.00269	.0187	.00006	.0013	.00376	.00159
%RSD	16.650	37.521	.46951	8.6416	.08919	10.706	52.853
#1	-.00201	-.00583	3.9724	.00069	1.4035	-.03701	-.00329
#2	-.00234	-.01026	3.9969	.00078	1.4046	-.03756	-.00130
#3	-.00167	-.00541	4.0093	.00066	1.4021	-.03079	-.00446
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						36.000	
Low Limit						-.03000	

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.00011**	**.03959**	**.00096**
Stddev	.00028	.00012	.00361
%RSD	248.38	.29182	376.14
#1	.00034	.03947	.00158
#2	-.00020	.03969	.00421
#3	.00020	.03962	-.00292
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9451.5**	**57637.**	**3009.6**
Stddev	22.8	186.	22.0
%RSD	.24088	.32311	.72999
#1	9477.8	57852.	2986.4
#2	9439.3	57520.	3012.4
#3	9437.5	57540.	3030.0

Approved: June 27, 2016

Sample Name: L1606098605 Acquired: 6/25/2016 19:55:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00137	.11904	-.00057	.01716	.07339	.00069	79.533	.00007
Stddev	.00111	.00237	.00183	.00141	.00107	.00010	.468	.00014
%RSD	81.063	1.9908	322.69	8.2448	1.4582	14.460	.58849	215.58
#1	-.00160	.12158	.00130	.01581	.07215	.00080	78.994	-.00002
#2	-.00016	.11865	-.00237	.01703	.07401	.00066	79.770	-.00001
#3	-.00235	.11689	-.00063	.01863	.07401	.00061	79.836	.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00018	.00197	.03453	.22031	1.0996	.00600	23.636	.02412
Stddev	.00010	.00062	.00164	.01719	.0321	.00378	.124	.00114
%RSD	58.539	31.534	4.7611	7.8016	2.9198	62.984	.52590	4.7055
#1	.00028	.00148	.03642	.22738	1.1284	.00174	23.677	.02484
#2	.00008	.00176	.03343	.23283	1.0650	.00731	23.734	.02281
#3	.00016	.00266	.03374	.20072	1.1054	.00895	23.496	.02471
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00088	16.543	.00139	.02145	.00230	-.00291	.00287	2.5906
Stddev	.00010	.077	.00029	.00514	.00167	.00113	.00378	.0043
%RSD	11.067	.46265	20.622	23.950	72.689	38.933	132.03	.16499
#1	.00094	16.521	.00147	.02583	.00422	-.00367	.00465	2.5950
#2	.00093	16.628	.00163	.02271	.00152	-.00161	-.00148	2.5865
#3	.00076	16.480	.00107	.01580	.00116	-.00345	.00543	2.5902
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098605 Acquired: 6/25/2016 19:55:15 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00022**	**.41216**	**-.00457**	**-.00401**	**.00095**	**.08292**	**.00137**
Stddev	.00045	.00253	.00113	.00092	.00095	.00025	.00112
%RSD	198.91	.61368	24.692	22.922	99.932	.30169	82.337
#1	.00020	.40988	-.00328	-.00486	-.00009	.08312	.00008
#2	.00068	.41488	-.00504	-.00303	.00116	.08301	.00218
#3	-.00021	.41173	-.00539	-.00414	.00177	.08264	.00183
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9691.4**	**59019.**	**3013.1**
Stddev	15.7	113.	12.9
%RSD	.16185	.19156	.42939
#1	9673.9	58889.	3028.0
#2	9704.2	59073.	3004.1
#3	9696.2	59095.	3007.4

Approved: June 27, 2016

Sample Name: L1606098607 Acquired: 6/25/2016 19:59:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00078	.03854	.00139	.01595	.03639	.00021	86.351	.00013
Stddev	.00196	.00438	.00171	.00299	.00069	.00005	.183	.00018
%RSD	252.38	11.356	122.54	18.743	1.8946	25.559	.21179	130.52
#1	-.00064	.04027	.00252	.01429	.03564	.00026	86.561	.00033
#2	-.00004	.03356	-.00057	.01941	.03654	.00022	86.226	.00007
#3	.00302	.04179	.00223	.01417	.03700	.00015	86.267	-.00000
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00086	.00096	.00901	.09595	1.3588	.01896	23.947	.13765
Stddev	.00012	.00155	.00054	.00336	.0905	.00285	.152	.00156
%RSD	14.271	161.05	6.0176	3.5057	6.6619	15.032	.63449	1.1314
#1	.00074	.00257	.00840	.09606	1.2544	.01772	23.782	.13585
#2	.00084	-.00052	.00942	.09926	1.4148	.02223	24.081	.13859
#3	.00099	.00083	.00922	.09254	1.4072	.01695	23.978	.13851
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	12.868	.00006	.00776	.00294	-.00141	.00160	3.6487
Stddev	.00036	.033	.00069	.00647	.00245	.00292	.00413	.0140
%RSD	74.294	.25857	1239.7	83.320	83.569	207.44	258.10	.38292
#1	.00008	12.905	-.00031	.01172	.00013	.00158	.00626	3.6621
#2	.00078	12.859	.00085	.01126	.00466	-.00154	-.00162	3.6496
#3	.00059	12.840	-.00038	.00030	.00403	-.00426	.00017	3.6342
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098607 Acquired: 6/25/2016 19:59:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**.85925**	**-.01297**	**-.00261**	**.00027**	**.01872**	**-.00023**
Stddev	.00058	.00368	.00482	.00118	.00049	.00005	.00100
%RSD	1135.5	.42821	37.150	45.319	182.28	.27976	440.95
#1	.00028	.86236	-.00802	-.00398	-.00019	.01878	.00007
#2	-.00072	.86019	-.01323	-.00183	.00021	.01871	.00059
#3	.00028	.85519	-.01765	-.00203	.00079	.01868	-.00134
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9667.2**	**58651.**	**3007.8**
Stddev	38.0	90.	17.4
%RSD	.39269	.15372	.57760
#1	9636.2	58610.	2989.6
#2	9655.8	58589.	3009.4
#3	9709.5	58755.	3024.3

Approved: June 27, 2016

Sample Name: L1606098607PS Acquired: 6/25/2016 20:03:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG574215-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19549**	**4.9683**	**.19699**	**.95431**	**.52192**	**.02489**	**83.384**	**.02448**
Stddev	.00145	.0012	.00037	.00069	.00335	.00004	.351	.00015
%RSD	.74419	.02324	.18968	.07193	.64144	.17436	.42141	.59448
#1	.19431	4.9671	.19669	.95367	.51889	.02485	83.062	.02455
#2	.19712	4.9693	.19740	.95504	.52551	.02494	83.759	.02458
#3	.19505	4.9687	.19686	.95422	.52137	.02490	83.330	.02431
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09612**	**.24445**	**.24901**	**2.0095**	**25.021**	**.52274**	**26.551**	**.36466**
Stddev	.00060	.00204	.00132	.0079	.209	.00262	.211	.00375
%RSD	.62715	.83504	.52860	.39401	.83514	.50106	.79467	1.0287
#1	.09620	.24322	.24870	2.0100	24.845	.52124	26.458	.36188
#2	.09668	.24681	.25046	2.0172	25.252	.52577	26.792	.36317
#3	.09548	.24332	.24788	2.0014	24.965	.52123	26.403	.36893
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50539**	**35.868**	**.24056**	**4.8300**	**.24330**	**.58213**	**.19141**	**5.6770**
Stddev	.00256	.121	.00174	.0165	.00225	.00117	.00102	.0113
%RSD	.50685	.33731	.72421	.34266	.92572	.20086	.53487	.19978
#1	.50625	35.730	.24173	4.8368	.24292	.58281	.19254	5.6837
#2	.50742	35.956	.24138	4.8421	.24572	.58078	.19056	5.6833
#3	.50251	35.919	.23856	4.8112	.24127	.58280	.19112	5.6639
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098607PS Acquired: 6/25/2016 20:03:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG574215-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48483**	**1.2623**	**.48068**	**.24615**	**.50488**	**.48944**	**.00427**
Stddev	.00147	.0060	.00508	.00275	.00238	.00115	.00064
%RSD	.30392	.47569	1.0573	1.1154	.47121	.23489	15.068
#1	.48451	1.2555	.47807	.24920	.50216	.48999	.00389
#2	.48644	1.2668	.48654	.24388	.50587	.49021	.00501
#3	.48354	1.2646	.47744	.24537	.50660	.48812	.00390
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9610.5**	**58019.**	**3004.1**
Stddev	20.4	180.	8.2
%RSD	.21278	.31055	.27313
#1	9607.7	58202.	3005.9
#2	9591.6	58012.	2995.2
#3	9632.2	57842.	3011.3

Approved: June 27, 2016

Sample Name: L1606098607SDL Acquired: 6/25/2016 20:06:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG574215-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00055	.00334	.00170	.00445	.00847	.00009	17.139	.00008
Stddev	.00049	.00730	.00411	.00207	.00008	.00007	.249	.00023
%RSD	88.649	218.69	241.70	46.624	.94243	75.660	1.4519	290.72
#1	.00089	.00593	.00612	.00219	.00843	.00006	17.426	.00034
#2	-.00001	.00899	-.00200	.00488	.00843	.00004	17.001	-.00003
#3	.00078	-.00490	.00098	.00627	.00857	.00017	16.989	-.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00056	.00037	.00191	.01894	.32539	.00400	4.9234	.02615
Stddev	.00005	.00023	.00083	.00475	.10069	.00194	.1236	.00303
%RSD	9.1063	60.795	43.755	25.087	30.943	48.604	2.5107	11.568
#1	.00062	.00059	.00284	.02402	.21108	.00176	4.9976	.02622
#2	.00053	.00038	.00165	.01821	.40094	.00496	4.9918	.02914
#3	.00053	.00014	.00123	.01460	.36416	.00527	4.7807	.02309
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00031	2.5977	.00028	.00852	.00297	.00160	.00156	.72149
Stddev	.00021	.0380	.00123	.00414	.00293	.00068	.00366	.00246
%RSD	66.333	1.4636	438.92	48.637	98.645	42.261	235.25	.34147
#1	.00008	2.6376	-.00081	.01268	-.00028	.00113	-.00104	.72244
#2	.00048	2.5938	.00004	.00439	.00377	.00130	.00574	.71869
#3	.00037	2.5619	.00161	.00849	.00542	.00238	-.00003	.72333
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098607SDL Acquired: 6/25/2016 20:06:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG574215-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**.17255**	**.00033**	**-.00095**	**.00100**	**.00419**	**.00060**
Stddev	.00072	.00063	.00618	.00123	.00055	.00013	.00060
%RSD	1344.6	.36773	1857.8	129.31	55.083	2.9985	99.890
#1	-.00060	.17322	-.00395	-.00224	.00116	.00417	-.00009
#2	.00076	.17246	-.00247	-.00083	.00039	.00432	.00094
#3	-.00032	.17196	.00742	.00021	.00145	.00407	.00094
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10100.**	**61152.**	**3087.7**
Stddev	17.	221.	3.3
%RSD	.16887	.36180	.10673
#1	10112.	61407.	3088.4
#2	10080.	61005.	3090.6
#3	10107.	61046.	3084.1

Approved: June 27, 2016

Sample Name: L1606098607SDL Acquired: 6/25/2016 20:10:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG574215-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00016**	**-.00220**	**-.00032**	**-.00063**	**.00197**	**.00015**	**3.3784**	**-.00009**
Stddev	.00281	.00599	.00141	.00122	.00070	.00004	.0403	.00029
%RSD	1738.4	271.64	438.98	193.96	35.714	23.124	1.1918	310.09
#1	.00289	-.00340	.00101	-.00187	.00220	.00018	3.3321	-.00013
#2	-.00273	.00429	-.00181	-.00058	.00253	.00017	3.3977	-.00037
#3	.00033	-.00751	-.00017	.00056	.00118	.00011	3.4054	.00021
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00001**	**-.00057**	**.00039**	**.03496**	**.02799**	**.00189**	**.96584**	**.00502**
Stddev	.00041	.00109	.00122	.02559	.02173	.00316	.06365	.00248
%RSD	5673.0	191.57	314.78	73.189	77.636	167.39	6.5899	49.517
#1	.00017	-.00161	.00178	.00957	.03029	.00425	1.0133	.00340
#2	.00028	-.00067	-.00011	.06074	.04847	.00313	.99069	.00377
#3	-.00047	.00057	-.00050	.03458	.00520	-.00171	.89352	.00788
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00037**	**.48659**	**-.00001**	**.00389**	**.00194**	**-.00589**	**.00123**	**.14706**
Stddev	.00019	.02107	.00027	.00257	.00032	.00198	.00158	.00411
%RSD	50.403	4.3293	3257.2	66.021	16.675	33.653	128.28	2.7928
#1	-.00035	.50181	-.00003	.00472	.00177	-.00726	-.00048	.14268
#2	-.00056	.46255	.00027	.00101	.00173	-.00680	.00263	.15082
#3	-.00019	.49542	-.00027	.00595	.00231	-.00362	.00154	.14770
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098607SDL Acquired: 6/25/2016 20:10:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG574215-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00022**	**.03404**	**-.00096**	**.00109**	**.00053**	**.00095**	**-.00050**
Stddev	.00021	.00026	.00497	.00270	.00028	.00007	.00147
%RSD	93.819	.76060	519.10	248.53	53.226	6.9237	292.99
#1	-.00042	.03414	.00415	-.00145	.00059	.00088	.00090
#2	-.00001	.03374	-.00578	.00079	.00078	.00097	-.00204
#3	-.00022	.03423	-.00124	.00392	.00022	.00101	-.00038
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10132.**	**61150.**	**3069.8**
Stddev	20.	142.	13.8
%RSD	.19324	.23287	.44985
#1	10144.	61269.	3085.6
#2	10109.	61188.	3063.7
#3	10143.	60992.	3060.0

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 20:14:26 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.39742**	**10.037**	**.39212**	**.48448**	**.99926**	**.05007**	**9.9665**
Stddev	.00078	.015	.00234	.00170	.00686	.00012	.0639
%RSD	.19741	.14886	.59686	.35046	.68696	.24571	.64080
#1	.39794	10.020	.39189	.48304	1.0007	.05009	10.032
#2	.39780	10.040	.39457	.48635	.99181	.05018	9.9043
#3	.39651	10.049	.38991	.48406	1.0053	.04994	9.9633
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.04995**	**.19662**	**.49719**	**.50408**	**3.9345**	**48.425**	**1.0056**
Stddev	.00017	.00028	.00293	.00085	.0512	.336	.0066
%RSD	.33202	.14096	.58868	.16772	1.3013	.69474	.65876
#1	.05012	.19642	.49505	.50367	3.9195	48.601	1.0066
#2	.04993	.19694	.50053	.50505	3.8924	48.037	.99858
#3	.04979	.19651	.49600	.50351	3.9915	48.637	1.0117
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.6539**	**.49874**	**1.0057**	**49.161**	**.50061**	**9.8144**	**.49889**
Stddev	.1152	.00402	.0014	.263	.00048	.0158	.00214
%RSD	1.1928	.80691	.13850	.53437	.09607	.16108	.42961
#1	9.7007	.49641	1.0053	49.311	.50117	9.7964	.49662
#2	9.5228	.49643	1.0072	48.858	.50038	9.8207	.49919
#3	9.7383	.50339	1.0045	49.316	.50029	9.8261	.50087
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 20:14:26 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1787**	**.38972**	**5.0552**	**.98705**	**.99934**	**.99527**	**.50770**
Stddev	.0014	.00525	.0102	.00133	.00573	.00628	.00313
%RSD	.12146	1.3477	.20130	.13498	.57306	.63102	.61635
#1	1.1790	.38370	5.0434	.98688	1.0013	.99951	.50582
#2	1.1772	.39212	5.0610	.98581	.99288	.98805	.50598
#3	1.1800	.39334	5.0611	.98846	1.0038	.99824	.51132
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0147**	**.99253**	F **2.9496**
Stddev	.0017	.00087	.0127
%RSD	.16286	.08813	.43029
#1	1.0131	.99259	2.9390
#2	1.0147	.99336	2.9462
#3	1.0164	.99162	2.9637
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9868.5**	**58806.**	**2964.6**
Stddev	19.7	151.	1.8
%RSD	.19965	.25680	.06126
#1	9885.2	58672.	2966.0
#2	9873.5	58776.	2965.4
#3	9846.8	58969.	2962.6

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 20:18:03 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00132**	**-.00340**	**-.00049**	**.00031**	**.00192**	**.00008**	**-.02002**	**-.00008**
Stddev	.00216	.00700	.00114	.00069	.00099	.00015	.01568	.00025
%RSD	163.88	205.77	232.07	221.43	51.768	179.89	78.343	317.94
#1	.00112	.00138	-.00179	.00077	.00240	-.00004	-.03265	.00011
#2	.00357	-.00014	-.00002	.00066	.00258	.00004	-.00247	-.00036
#3	-.00074	-.01144	.00034	-.00049	.00078	.00025	-.02493	.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00074**	**.00038**	**.00063**	**.02332**	**-.07555**	**-.00123**	**.02786**	**-.00147**
Stddev	.00036	.00095	.00188	.02448	.09270	.00682	.09570	.00329
%RSD	47.976	247.76	298.28	104.94	122.70	553.52	343.47	223.66
#1	.00080	.00146	-.00061	.04476	-.06611	.00634	-.06528	.00195
#2	.00036	.00002	.00279	-.00335	-.17262	-.00687	.02294	-.00462
#3	.00107	-.00033	-.00029	.02856	.01207	-.00317	.12594	-.00174
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00377**	**.03770**	**.00015**	**.00241**	**-.00006**	**.00325**	**.00675**	**.03903**
Stddev	.00019	.01915	.00084	.00431	.00100	.00269	.00806	.00132
%RSD	5.0202	50.799	576.30	178.78	1573.6	82.788	119.45	3.3700
#1	.00388	.01718	.00046	.00371	.00063	.00040	.01512	.04015
#2	.00389	.04082	.00078	-.00240	-.00121	.00574	.00608	.03936
#3	.00355	.05510	-.00080	.00592	.00039	.00360	-.00096	.03758
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 20:18:03 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00001**	**.00027**	**.00088**	**-.00106**	**.00004**	**-.00003**	**.00449**
Stddev	.00046	.00064	.00516	.00225	.00089	.00028	.00124
%RSD	6827.2	239.61	583.41	212.27	2439.5	892.40	27.555
#1	.00053	-.00043	-.00417	-.00308	-.00098	.00021	.00367
#2	-.00031	.00040	.00068	.00137	.00041	.00004	.00389
#3	-.00023	.00082	.00615	-.00147	.00068	-.00034	.00592
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9631.9**	**58363.**	**2969.7**
Stddev	4.5	93.	16.5
%RSD	.04685	.16010	.55454
#1	9627.4	58461.	2965.8
#2	9636.4	58275.	2987.8
#3	9631.8	58353.	2955.6

Approved: June 27, 2016

Sample Name: L1606098609 Acquired: 6/25/2016 20:22:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00081	.68625	.00096	.00637	.02063	.00019	116.68	.00011
Stddev	.00130	.00243	.00203	.00124	.00115	.00006	.34	.00020
%RSD	161.33	.35410	210.74	19.489	5.5807	30.863	.28955	179.93
#1	.00013	.68352	.00323	.00768	.02147	.00021	116.54	.00014
#2	.00231	.68708	-.00070	.00521	.01932	.00023	116.43	-.00010
#3	-.00002	.68816	.00037	.00620	.02109	.00012	117.06	.00029
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	.00251	.00589	.80058	.71511	.00736	28.613	.02306
Stddev	.00014	.00088	.00100	.02799	.18771	.00368	.079	.00071
%RSD	23.262	34.910	17.057	3.4966	26.250	49.956	.27754	3.0949
#1	.00051	.00150	.00680	.80914	.91494	.01159	28.526	.02382
#2	.00075	.00293	.00481	.76931	.68792	.00556	28.632	.02293
#3	.00052	.00310	.00605	.82330	.54247	.00492	28.681	.02241
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00074	14.713	.00303	.03333	.00106	-.00272	.00365	4.8889
Stddev	.00024	.075	.00054	.00378	.00114	.00271	.00111	.0101
%RSD	32.746	.50774	17.920	11.335	108.20	99.808	30.510	.20564
#1	.00047	14.684	.00299	.03128	.00107	-.00577	.00444	4.8999
#2	.00095	14.656	.00359	.03769	-.00009	-.00056	.00238	4.8868
#3	.00081	14.797	.00251	.03102	.00220	-.00183	.00413	4.8801
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098609 Acquired: 6/25/2016 20:22:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00003	.51888	.00260	-.00246	.00196	.03426	.00729
Stddev	.00032	.00275	.00453	.00099	.00037	.00011	.00028
%RSD	1222.6	.53076	174.15	40.180	18.856	.31556	3.8509
#1	.00009	.51844	-.00256	-.00245	.00220	.03417	.00723
#2	-.00032	.51637	.00591	-.00346	.00153	.03438	.00705
#3	.00031	.52183	.00446	-.00148	.00215	.03423	.00760
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9797.7	59515.	3093.6
Stddev	11.0	17.	11.2
%RSD	.11267	.02773	.36079
#1	9791.0	59496.	3083.8
#2	9810.4	59520.	3105.7
#3	9791.6	59528.	3091.1

Approved: June 27, 2016

Sample Name: L1606098611 Acquired: 6/25/2016 20:25:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00025	.67808	-.00318	.08715	.06918	.00114	148.14	.00017
Stddev	.00037	.00569	.00382	.00198	.00070	.00012	.69	.00021
%RSD	151.84	.83972	120.10	2.2728	1.0151	10.946	.46623	125.05
#1	-.00014	.67215	-.00172	.08491	.06933	.00099	148.49	.00006
#2	.00061	.68350	-.00031	.08867	.06980	.00119	147.34	.00003
#3	.00027	.67861	-.00751	.08788	.06842	.00123	148.59	.00041
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00363	.00334	.07531	1.1555	1.9378	.02263	43.443	.16489
Stddev	.00025	.00064	.00057	.0192	.0114	.00045	.331	.00288
%RSD	6.9167	19.099	.75123	1.6613	.58882	1.9873	.76282	1.7487
#1	.00388	.00297	.07511	1.1584	1.9507	.02262	43.610	.16408
#2	.00338	.00297	.07595	1.1731	1.9292	.02219	43.061	.16250
#3	.00362	.00407	.07487	1.1351	1.9335	.02309	43.657	.16809
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00248	25.675	.00454	.04285	.00455	-.00035	-.00286	3.7826
Stddev	.00024	.125	.00067	.00695	.00113	.00106	.00555	.0053
%RSD	9.7863	.48812	14.843	16.217	24.766	306.62	194.07	.13999
#1	.00260	25.775	.00423	.04366	.00585	.00053	-.00927	3.7791
#2	.00220	25.535	.00409	.03554	.00387	-.00004	.00020	3.7887
#3	.00264	25.716	.00532	.04936	.00394	-.00153	.00049	3.7799
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098611 Acquired: 6/25/2016 20:25:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00008	2.0876	-.00638	-.00280	.00142	.06566	.00394
Stddev	.00055	.0096	.00925	.00066	.00022	.00017	.00141
%RSD	713.59	.45905	144.86	23.607	15.199	.25178	35.771
#1	-.00017	2.0923	-.00754	-.00222	.00163	.06585	.00254
#2	-.00031	2.0765	-.01500	-.00266	.00143	.06556	.00536
#3	.00071	2.0939	.00339	-.00351	.00120	.06556	.00393
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9695.5	58902.	3054.3
Stddev	10.1	156.	10.7
%RSD	.10398	.26470	.35100
#1	9691.1	59006.	3048.2
#2	9688.3	58978.	3066.6
#3	9707.0	58723.	3048.0

Approved: June 27, 2016

Sample Name: L1606098613 Acquired: 6/25/2016 20:29:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00174	5.5730	.00432	.03801	.11035	.00130	203.09	-.00001
Stddev	.00149	.0359	.00264	.00141	.00031	.00003	.17	.00016
%RSD	85.559	.64338	61.135	3.7127	.27868	2.0129	.08522	1090.1
#1	.00010	5.6007	.00735	.03819	.11034	.00128	203.27	.00012
#2	.00211	5.5858	.00310	.03932	.11006	.00133	203.05	.00003
#3	.00302	5.5325	.00251	.03651	.11067	.00131	202.93	-.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00373	.00967	.08024	7.4517	4.1196	.03087	79.593	.34313
Stddev	.00008	.00046	.00072	.0194	.0319	.00148	.239	.00219
%RSD	2.0959	4.7129	.90070	.26051	.77317	4.7986	.30050	.63957
#1	.00364	.00917	.07941	7.4293	4.0890	.03222	79.370	.34059
#2	.00377	.01007	.08072	7.4635	4.1171	.03112	79.565	.34436
#3	.00379	.00978	.08059	7.4624	4.1526	.02929	79.846	.34443
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00182	24.658	.00975	.17618	.00825	.00008	.00248	10.141
Stddev	.00040	.019	.00024	.00598	.00295	.00311	.00271	.073
%RSD	21.744	.07518	2.5015	3.3961	35.754	3886.1	109.35	.72041
#1	.00146	24.670	.00990	.17816	.00502	.00353	.00373	10.058
#2	.00175	24.669	.00947	.16946	.00891	-.00250	.00433	10.169
#3	.00224	24.637	.00989	.18093	.01080	-.00079	-.00063	10.196
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098613 Acquired: 6/25/2016 20:29:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00066	2.4743	.05397	-.00080	.01073	.41856	.00624
Stddev	.00027	.0052	.00665	.00236	.00131	.00229	.00115
%RSD	39.960	.21077	12.313	294.11	12.169	.54671	18.419
#1	.00089	2.4751	.05254	-.00353	.01217	.41596	.00756
#2	.00073	2.4791	.04816	.00063	.00962	.42027	.00548
#3	.00037	2.4687	.06122	.00049	.01040	.41945	.00567
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9619.4	58503.	3058.9
Stddev	38.3	135.	15.7
%RSD	.39844	.23102	.51315
#1	9576.0	58658.	3041.4
#2	9633.7	58414.	3071.7
#3	9648.5	58435.	3063.7

Approved: June 27, 2016

Sample Name: L1606098615 Acquired: 6/25/2016 20:33:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00201	.53689	.00109	.12769	.02932	.00091	261.33	.00014
Stddev	.00048	.01300	.00276	.00142	.00104	.00011	1.37	.00008
%RSD	23.740	2.4215	252.93	1.1120	3.5600	12.063	.52322	55.108
#1	.00148	.54532	.00299	.12707	.02978	.00081	261.04	.00021
#2	.00215	.54344	.00237	.12931	.03005	.00088	262.83	.00006
#3	.00240	.52192	-.00208	.12669	.02812	.00103	260.14	.00015
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00075	.00167	.04808	1.7755	3.9837	.04482	96.121	.35461
Stddev	.00009	.00172	.00105	.0329	.0639	.00305	.330	.00305
%RSD	12.352	103.13	2.1885	1.8532	1.6044	6.8091	.34352	.86008
#1	.00068	.00125	.04714	1.8121	3.9806	.04242	96.172	.35269
#2	.00071	.00356	.04922	1.7483	4.0491	.04379	96.422	.35813
#3	.00085	.00020	.04788	1.7662	3.9214	.04825	95.768	.35302
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00080	71.586	.00152	.06528	.00543	-.00023	-.00042	6.3830
Stddev	.00049	.359	.00083	.00282	.00224	.00292	.00309	.0073
%RSD	60.493	.50193	55.060	4.3183	41.272	1275.7	729.65	.11354
#1	.00036	71.588	.00097	.06827	.00795	-.00316	.00167	6.3913
#2	.00132	71.945	.00248	.06488	.00468	-.00019	.00103	6.3782
#3	.00072	71.227	.00110	.06267	.00366	.00267	-.00398	6.3795
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098615 Acquired: 6/25/2016 20:33:32 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00066**	**6.9477**	**-.01868**	**-.00208**	**-.00036**	**.07416**	**.00363**
Stddev	.00094	.0407	.00365	.00070	.00044	.00028	.00062
%RSD	142.12	.58617	19.553	33.577	122.09	.37861	17.066
#1	.00171	6.9428	-.02024	-.00215	-.00034	.07441	.00412
#2	.00039	6.9907	-.01451	-.00134	-.00081	.07386	.00293
#3	-.00011	6.9097	-.02130	-.00273	.00007	.07420	.00382
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9482.1**	**57521.**	**3041.1**
Stddev	15.4	119.	13.0
%RSD	.16231	.20772	.42869
#1	9472.3	57649.	3053.5
#2	9499.8	57500.	3027.5
#3	9474.1	57412.	3042.2

Approved: June 27, 2016

Sample Name: L1606098617 Acquired: 6/25/2016 20:37:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00154	.20547	.00214	.04863	.05628	.00015	62.798	.00005
Stddev	.00260	.00707	.00399	.00222	.00076	.00003	.506	.00026
%RSD	168.84	3.4433	186.39	4.5608	1.3548	21.842	.80652	579.39
#1	.00162	.19836	-.00117	.04752	.05710	.00015	63.380	-.00016
#2	.00410	.20553	.00657	.05118	.05613	.00018	62.459	.00034
#3	-.00110	.21251	.00102	.04718	.05560	.00012	62.555	-.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00034	.00139	.00754	.49751	2.0420	.02056	20.601	.01571
Stddev	.00024	.00096	.00096	.02565	.1117	.00244	.034	.00089
%RSD	70.093	68.878	12.744	5.1552	5.4689	11.879	.16418	5.6341
#1	.00059	.00246	.00739	.47113	2.0254	.02185	20.628	.01672
#2	.00012	.00112	.00857	.52236	1.9395	.01774	20.613	.01507
#3	.00031	.00060	.00666	.49904	2.1610	.02209	20.563	.01535
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00090	26.598	.00080	.00703	.00246	-.00325	.00131	3.4906
Stddev	.00064	.178	.00058	.00500	.00246	.00605	.00554	.0173
%RSD	70.655	.66988	72.589	71.163	99.983	186.20	421.53	.49585
#1	.00155	26.804	.00039	.00280	.00327	-.00237	-.00267	3.5042
#2	.00088	26.495	.00146	.00574	-.00030	-.00969	.00764	3.4966
#3	.00027	26.495	.00054	.01256	.00441	.00231	-.00103	3.4711
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098617 Acquired: 6/25/2016 20:37:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00032	1.5150	-.00311	-.00236	-.00015	.01447	.00341
Stddev	.00072	.0088	.00641	.00068	.00005	.00011	.00060
%RSD	226.25	.58268	206.23	28.717	34.447	.73139	17.615
#1	.00066	1.5251	-.00263	-.00291	-.00009	.01435	.00287
#2	.00080	1.5088	.00305	-.00160	-.00015	.01454	.00406
#3	-.00051	1.5111	-.00975	-.00256	-.00020	.01452	.00331
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9830.7	59822.	3082.9
Stddev	15.5	147.	23.6
%RSD	.15717	.24641	.76693
#1	9813.6	59963.	3057.5
#2	9843.5	59835.	3104.4
#3	9835.1	59669.	3086.7

Approved: June 27, 2016

Sample Name: L1606098619 Acquired: 6/25/2016 20:41:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00113	.27749	.00069	.06930	.05323	.00039	10.214	.00034
Stddev	.00263	.00318	.00178	.00081	.00134	.00007	.065	.00038
%RSD	232.09	1.1459	258.13	1.1649	2.5186	17.965	.63738	111.28
#1	.00355	.27762	.00105	.06852	.05477	.00042	10.240	.00066
#2	-.00167	.28061	.00227	.06924	.05251	.00044	10.140	.00044
#3	.00153	.27425	-.00125	.07014	.05240	.00031	10.262	-.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00108	.00071	.02585	.24748	1.2811	.01465	4.6162	.00999
Stddev	.00051	.00114	.00067	.00474	.0573	.00378	.0318	.00359
%RSD	46.988	159.40	2.5887	1.9167	4.4759	25.817	.68816	35.904
#1	.00062	.00202	.02645	.25048	1.2371	.01153	4.5886	.01278
#2	.00162	-.00003	.02513	.24201	1.3460	.01356	4.6509	.01123
#3	.00099	.00015	.02599	.24994	1.2604	.01885	4.6090	.00594
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00123	228.52	.00044	.07963	.00371	-.00207	.00619	2.5103
Stddev	.00028	1.82	.00063	.00808	.00266	.00077	.00393	.0173
%RSD	22.669	.79720	141.35	10.151	71.562	37.379	63.490	.68709
#1	.00153	229.52	.00003	.07034	.00201	-.00126	.00175	2.5018
#2	.00118	226.42	.00013	.08504	.00235	-.00280	.00758	2.4989
#3	.00097	229.62	.00117	.08351	.00677	-.00216	.00924	2.5301
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098619 Acquired: 6/25/2016 20:41:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00007	.22967	.00683	-.00264	.00219	.02256	.00408
Stddev	.00091	.00074	.00628	.00047	.00030	.00021	.00266
%RSD	1371.8	.32350	91.943	17.727	13.766	.92425	65.101
#1	-.00024	.23041	.00516	-.00255	.00222	.02259	.00112
#2	-.00065	.22892	.00155	-.00314	.00188	.02235	.00625
#3	.00109	.22967	.01377	-.00222	.00248	.02276	.00487
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9702.8	58480.	3067.3
Stddev	45.7	28.	15.6
%RSD	.47062	.04775	.50880
#1	9725.9	58510.	3053.5
#2	9732.4	58474.	3084.2
#3	9650.3	58455.	3064.3

Approved: June 27, 2016

Sample Name: L1606098621 Acquired: 6/25/2016 20:45:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00057	.00121	.00179	.05746	.02986	.00029	128.59	-.00000
Stddev	.00056	.00275	.00384	.00358	.00132	.00008	.53	.00025
%RSD	97.573	226.34	214.60	6.2318	4.4201	27.216	.41011	6271.6
#1	.00040	-.00100	.00602	.06157	.03100	.00034	129.12	-.00029
#2	.00120	.00036	-.00148	.05576	.03017	.00020	128.07	.00019
#3	.00012	.00429	.00083	.05504	.02841	.00032	128.57	.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00019	.00173	.01434	.05981	2.5098	.02550	33.269	.15420
Stddev	.00017	.00097	.00185	.00891	.0950	.00146	.060	.00170
%RSD	92.832	55.806	12.905	14.892	3.7837	5.7272	.18077	1.1012
#1	.00010	.00269	.01260	.05052	2.5555	.02390	33.334	.15465
#2	.00007	.00177	.01629	.06064	2.4007	.02586	33.216	.15563
#3	.00038	.00075	.01412	.06828	2.5733	.02675	33.257	.15232
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00056	47.893	-.00000	.01004	.00162	-.00276	-.00417	3.5790
Stddev	.00060	.257	.00189	.00801	.00056	.00235	.00172	.0047
%RSD	106.13	.53712	269880.	79.763	34.469	85.173	41.117	.13102
#1	.00071	48.149	.00182	.00080	.00218	-.00406	-.00514	3.5827
#2	-.00009	47.634	.00013	.01426	.00163	-.00417	-.00519	3.5737
#3	.00107	47.897	-.00195	.01505	.00106	-.00005	-.00219	3.5806
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098621 Acquired: 6/25/2016 20:45:10 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00024**	**1.8050**	**-.02231**	**-.00291**	**.00001**	**.01004**	**.00207**
Stddev	.00095	.0120	.00564	.00294	.00125	.00002	.00101
%RSD	393.73	.66689	25.288	100.88	21161.	.16952	48.549
#1	.00057	1.8173	-.02371	-.00622	-.00119	.01005	.00218
#2	-.00128	1.7932	-.02711	-.00189	.00130	.01004	.00102
#3	-.00001	1.8046	-.01610	-.00062	-.00009	.01002	.00302
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9662.5**	**58180.**	**3042.4**
Stddev	14.0	47.	7.3
%RSD	.14521	.08141	.24078
#1	9658.0	58146.	3035.1
#2	9678.2	58161.	3042.4
#3	9651.2	58234.	3049.8

Approved: June 27, 2016

Sample Name: L1606098623 Acquired: 6/25/2016 20:49:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00139	.10705	.00198	.41392	.02623	.00027	243.17	.00017
Stddev	.00218	.00541	.00096	.00428	.00052	.00003	1.41	.00011
%RSD	157.63	5.0515	48.506	1.0332	1.9773	12.148	.58072	65.505
#1	.00177	.11044	.00295	.41225	.02682	.00026	243.56	.00005
#2	.00336	.10081	.00102	.41878	.02603	.00030	244.34	.00020
#3	-.00096	.10989	.00198	.41074	.02584	.00024	241.60	.00026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00008	.00125	.02256	.13395	6.1008	.01177	68.965	.45308
Stddev	.00021	.00090	.00053	.00942	.0335	.00473	.254	.00553
%RSD	267.67	72.031	2.3661	7.0298	.54942	40.196	.36871	1.2196
#1	-.00012	.00056	.02318	.12309	6.1028	.01598	68.820	.44996
#2	.00031	.00093	.02221	.13890	6.0663	.00665	69.259	.45946
#3	.00005	.00227	.02231	.13986	6.1333	.01267	68.816	.44983
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00044	66.667	.00075	.01360	.00015	-.00234	.00238	3.6345
Stddev	.00039	.343	.00118	.00476	.00283	.00395	.00376	.0082
%RSD	88.357	.51401	157.96	35.012	1938.3	168.93	158.02	.22548
#1	.00001	66.765	.00108	.01427	-.00296	-.00415	.00669	3.6345
#2	.00056	66.950	-.00056	.00853	.00082	.00219	.00057	3.6426
#3	.00076	66.286	.00172	.01798	.00259	-.00506	-.00014	3.6263
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098623 Acquired: 6/25/2016 20:49:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	1.6999	-.02956	-.00156	.00030	.00857	.00329
Stddev	.00028	.0081	.00507	.00230	.00048	.00011	.00135
%RSD	598.12	.47536	17.163	147.79	160.42	1.3327	41.067
#1	.00005	1.7017	-.03255	.00072	-.00015	.00844	.00173
#2	.00033	1.7069	-.02370	-.00388	.00023	.00859	.00400
#3	-.00023	1.6911	-.03242	-.00151	.00081	.00866	.00415
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9416.4	57429.	3035.8
Stddev	48.0	130.	25.1
%RSD	.51009	.22575	.82551
#1	9467.2	57482.	3016.2
#2	9410.3	57282.	3027.1
#3	9371.8	57525.	3064.1

Approved: June 27, 2016

Sample Name: L1606098625 Acquired: 6/25/2016 20:52:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00225	2.3612	.00034	.04147	.11975	.00023	121.59	.00037
Stddev	.00036	.0069	.00441	.00209	.00117	.00012	.33	.00010
%RSD	15.843	.29306	1297.6	5.0487	.97876	50.808	.26806	27.630
#1	.00211	2.3674	.00539	.03961	.11950	.00036	121.76	.00028
#2	.00199	2.3537	-.00273	.04106	.11872	.00014	121.21	.00033
#3	.00266	2.3625	-.00164	.04374	.12102	.00019	121.79	.00048
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00161	.00414	.00666	2.3953	2.5999	.01491	37.526	1.0679
Stddev	.00025	.00072	.00102	.0202	.0477	.00333	.332	.0124
%RSD	15.696	17.478	15.382	.84513	1.8346	22.308	.88561	1.1598
#1	.00188	.00385	.00549	2.3789	2.6109	.01808	37.493	1.0637
#2	.00138	.00496	.00714	2.4179	2.5476	.01145	37.212	1.0581
#3	.00156	.00361	.00736	2.3890	2.6411	.01519	37.874	1.0818
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00064	18.226	.00269	.05391	.00651	-.00190	.00505	5.9129
Stddev	.00020	.052	.00077	.00146	.00208	.00291	.00127	.0400
%RSD	30.881	.28797	28.786	2.7065	31.918	152.94	25.077	.67624
#1	.00067	18.245	.00228	.05412	.00658	-.00209	.00366	5.9121
#2	.00082	18.167	.00221	.05526	.00440	.00110	.00533	5.9533
#3	.00043	18.267	.00358	.05236	.00855	-.00472	.00614	5.8733
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098625 Acquired: 6/25/2016 20:52:55 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00035**	**.99462**	**.03044**	**-.00185**	**.00457**	**.15726**	**.00437**
Stddev	.00053	.00442	.00913	.00019	.00047	.00042	.00072
%RSD	150.91	.44485	29.990	10.187	10.282	.27017	16.565
#1	.00010	.99601	.03923	-.00207	.00503	.15763	.00393
#2	-.00093	.98967	.02100	-.00171	.00410	.15735	.00520
#3	-.00022	.99819	.03111	-.00178	.00457	.15680	.00397
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9574.3**	**58285.**	**3022.7**
Stddev	18.0	80.	21.3
%RSD	.18820	.13801	.70548
#1	9553.8	58363.	3008.3
#2	9581.8	58202.	3047.2
#3	9587.4	58291.	3012.6

Approved: June 27, 2016

Sample Name: L1606098627 Acquired: 6/25/2016 20:56:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00254	4.6384	.00407	.03588	.08487	.00132	200.73	.00005
Stddev	.00192	.0173	.00097	.00149	.00032	.00011	.32	.00023
%RSD	75.499	.37394	23.741	4.1641	.37451	8.4866	.15998	422.09
#1	.00186	4.6215	.00464	.03631	.08521	.00120	200.37	.00006
#2	.00106	4.6377	.00461	.03711	.08483	.00141	200.85	-.00018
#3	.00471	4.6561	.00295	.03422	.08457	.00136	200.98	.00028
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00295	.00847	.07661	6.3761	4.0467	.02688	78.393	.28594
Stddev	.00029	.00096	.00098	.0153	.0946	.00177	.290	.00214
%RSD	9.9470	11.337	1.2842	.24023	2.3374	6.5906	.37011	.74836
#1	.00263	.00958	.07761	6.3586	3.9534	.02860	78.140	.28832
#2	.00299	.00791	.07658	6.3825	4.0443	.02506	78.709	.28419
#3	.00322	.00791	.07564	6.3872	4.1425	.02698	78.329	.28530
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00106	25.068	.00713	.15173	.00678	-.00194	.00054	8.9965
Stddev	.00058	.104	.00038	.00131	.00344	.00137	.00531	.0286
%RSD	55.068	.41293	5.2660	.86550	50.818	70.752	976.81	.31838
#1	.00130	24.949	.00680	.15078	.01021	-.00340	-.00348	9.0111
#2	.00040	25.140	.00754	.15323	.00679	-.00068	-.00144	9.0150
#3	.00149	25.114	.00704	.15117	.00333	-.00174	.00656	8.9635
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098627 Acquired: 6/25/2016 20:56:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00089	2.4772	.04490	-.00127	.00901	.16031	.00422
Stddev	.00154	.0067	.00796	.00182	.00061	.00117	.00055
%RSD	173.10	.26965	17.717	142.95	6.7328	.72957	12.944
#1	.00142	2.4695	.04447	-.00241	.00911	.16067	.00359
#2	.00210	2.4809	.03717	.00082	.00836	.16126	.00459
#3	-.00085	2.4812	.05306	-.00223	.00956	.15900	.00448
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9465.6	57841.	2995.1
Stddev	68.7	160.	17.6
%RSD	.72586	.27642	.58883
#1	9445.1	57701.	3013.7
#2	9409.5	58015.	2978.6
#3	9542.2	57807.	2993.1

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 21:00:36 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.40235	10.185	.39842	.49195	1.0203	.05034	10.153
Stddev	.00325	.028	.00193	.00194	.0055	.00031	.030
%RSD	.80675	.27855	.48505	.39452	.53912	.62290	.29151
#1	.39985	10.214	.39899	.49235	1.0140	.05068	10.126
#2	.40602	10.183	.40000	.49365	1.0227	.05026	10.185
#3	.40118	10.157	.39627	.48984	1.0242	.05007	10.150
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05123	.19988	.50288	.51330	4.0070	49.596	1.0284
Stddev	.00026	.00047	.00335	.00193	.0033	.230	.0041
%RSD	.51360	.23712	.66679	.37527	.08271	.46359	.40302
#1	.05102	.20030	.50672	.51551	4.0072	49.649	1.0258
#2	.05153	.19997	.50051	.51234	4.0102	49.344	1.0261
#3	.05115	.19937	.50141	.51204	4.0036	49.795	1.0332
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.5669	.50781	1.0249	50.362	.50842	9.9814	.50557
Stddev	.0485	.00742	.0065	.234	.00100	.0181	.00207
%RSD	.50657	1.4609	.62973	.46520	.19680	.18135	.40980
#1	9.6226	.50131	1.0311	50.266	.50877	9.9736	.50596
#2	9.5340	.50622	1.0255	50.191	.50920	10.002	.50742
#3	9.5442	.51589	1.0183	50.629	.50729	9.9685	.50333
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 21:00:36 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2045**	**.39202**	**5.1444**	**1.0010**	**1.0190**	**1.0190**	**.51206**
Stddev	.0033	.00153	.0072	.0031	.0045	.0084	.00251
%RSD	.27464	.38961	.14055	.30711	.44063	.82168	.49020
#1	1.2010	.39059	5.1361	1.0039	1.0167	1.0152	.51490
#2	1.2047	.39363	5.1492	1.0015	1.0162	1.0132	.51015
#3	1.2076	.39184	5.1479	.99778	1.0242	1.0286	.51112
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0223**	**1.0141**	F **2.9929**
Stddev	.0031	.0001	.0241
%RSD	.30006	.01208	.80650
#1	1.0229	1.0141	3.0133
#2	1.0190	1.0142	2.9663
#3	1.0251	1.0139	2.9992
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9649.2**	**58649.**	**2955.6**
Stddev	2.7	54.	16.9
%RSD	.02842	.09258	.57071
#1	9646.1	58667.	2944.4
#2	9650.3	58588.	2975.0
#3	9651.2	58692.	2947.5

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 21:04:14 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00098**	**-.00081**	**.00091**	**.00315**	**.00292**	**.00009**	**-.03474**	**.00002**
Stddev	.00124	.00864	.00092	.00201	.00133	.00005	.02919	.00008
%RSD	125.96	1064.0	100.48	64.019	45.548	52.705	84.022	417.72
#1	.00043	-.00670	.00067	.00330	.00336	.00005	-.02620	.00002
#2	-.00151	.00910	.00014	.00508	.00142	.00009	-.01077	-.00006
#3	-.00188	-.00483	.00193	.00106	.00397	.00014	-.06724	.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00004**	**.00125**	**.00083**	**.02088**	**-.14334**	**.00212**	**-.02955**	**.00114**
Stddev	.00066	.00051	.00122	.01391	.04593	.00105	.09306	.00075
%RSD	1530.9	41.168	147.43	66.595	32.046	49.608	314.87	66.235
#1	.00067	.00069	.00144	.03293	-.09030	.00098	-.05631	.00047
#2	-.00016	.00170	-.00058	.02406	-.17011	.00232	-.10630	.00196
#3	-.00064	.00135	.00162	.00566	-.16960	.00305	.07395	.00100
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00323**	**.05080**	**-.00020**	**.00483**	**.00058**	**.00191**	**.00111**	**.03921**
Stddev	.00021	.02575	.00089	.00353	.00235	.00387	.00242	.00153
%RSD	6.4882	50.676	441.28	73.104	406.34	202.21	216.78	3.9059
#1	.00300	.07783	.00045	.00259	-.00201	.00626	-.00135	.03746
#2	.00339	.02657	-.00122	.00300	.00257	.00062	.00121	.03991
#3	.00331	.04802	.00016	.00889	.00117	-.00114	.00348	.04027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 21:04:14 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	.00029	-.00007	-.00513	.00041	-.00001	.00470
Stddev	.00027	.00000	.00626	.00369	.00015	.00012	.00055
%RSD	55.343	1.4693	8856.4	71.949	37.134	916.05	11.589
#1	.00081	.00028	.00338	-.00666	.00031	.00010	.00427
#2	.00033	.00029	-.00730	-.00092	.00059	.00000	.00453
#3	.00034	.00028	.00370	-.00780	.00035	-.00014	.00531
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9506.6	58497.	2929.7
Stddev	22.3	113.	12.0
%RSD	.23433	.19372	.40841
#1	9532.3	58628.	2919.5
#2	9494.8	58435.	2926.6
#3	9492.6	58429.	2942.8

Approved: June 27, 2016

Sample Name: L1606098629 Acquired: 6/25/2016 21:08:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00070	1.3154	-.00179	.12516	.05800	.00081	247.18	.00018
Stddev	.00097	.0064	.00105	.00195	.00108	.00003	.31	.00008
%RSD	137.44	.48564	58.530	1.5550	1.8687	3.6829	.12350	43.911
#1	.00160	1.3100	-.00292	.12302	.05810	.00083	246.92	.00018
#2	.00082	1.3224	-.00085	.12682	.05904	.00078	247.52	.00010
#3	-.00032	1.3138	-.00160	.12563	.05687	.00083	247.09	.00026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00119	.00401	.04528	2.4364	3.9714	.04409	89.669	.34274
Stddev	.00048	.00063	.00072	.0178	.1152	.00255	.458	.00095
%RSD	40.012	15.569	1.5917	.73079	2.9014	5.7752	.51098	.27642
#1	.00152	.00449	.04490	2.4275	3.9851	.04702	90.007	.34368
#2	.00141	.00424	.04612	2.4569	4.0792	.04281	89.852	.34276
#3	.00065	.00331	.04484	2.4248	3.8500	.04243	89.147	.34178
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00084	72.159	.00306	.12842	.00472	-.00245	-.00232	7.1068
Stddev	.00019	.103	.00116	.00307	.00202	.00283	.00582	.0122
%RSD	22.591	.14315	38.091	2.3898	42.715	115.64	250.20	.17182
#1	.00062	72.124	.00297	.12975	.00386	-.00541	-.00257	7.1178
#2	.00090	72.275	.00194	.13059	.00327	-.00216	-.00801	7.1089
#3	.00099	72.077	.00426	.12491	.00702	.00023	.00361	7.0937
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098629 Acquired: 6/25/2016 21:08:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00063**	**6.8039**	**.00261**	**-.00311**	**.00203**	**.14656**	**.00713**
Stddev	.00109	.0140	.00201	.00153	.00039	.00030	.00139
%RSD	174.70	.20570	77.238	49.135	18.962	.20399	19.505
#1	.00026	6.7979	.00196	-.00298	.00228	.14634	.00715
#2	-.00024	6.8199	.00487	-.00165	.00159	.14690	.00573
#3	.00185	6.7939	.00100	-.00469	.00223	.14643	.00851
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9470.9**	**57988.**	**3007.5**
Stddev	19.0	71.	5.1
%RSD	.20079	.12320	.16890
#1	9472.3	57910.	3001.8
#2	9451.2	58005.	3011.5
#3	9489.2	58049.	3009.1

Approved: June 27, 2016

Sample Name: L1606098631S Acquired: 6/25/2016 21:12:02 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.18827	4.8059	.19457	.97775	.51082	.02368	136.28	.02401
Stddev	.00146	.0060	.00214	.00486	.00158	.00008	.55	.00011
%RSD	.77669	.12402	1.0980	.49690	.30853	.31910	.40245	.44044
#1	.18949	4.8104	.19700	.98276	.50954	.02372	135.66	.02402
#2	.18665	4.7991	.19298	.97305	.51258	.02359	136.67	.02389
#3	.18867	4.8083	.19374	.97744	.51033	.02373	136.53	.02410
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09298	.23341	.25035	1.9823	26.282	.52091	38.371	.48555
Stddev	.00063	.00083	.00168	.0210	.198	.00132	.105	.00060
%RSD	.67753	.35580	.67127	1.0601	.75241	.25420	.27268	.12393
#1	.09358	.23411	.25193	1.9586	26.065	.52066	38.253	.48489
#2	.09232	.23249	.25054	1.9895	26.453	.52234	38.408	.48570
#3	.09303	.23364	.24858	1.9987	26.329	.51972	38.452	.48607
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.49539	77.049	.23394	4.7766	.23491	.57914	.18644	6.1257
Stddev	.00233	.345	.00087	.0076	.00312	.00333	.00623	.0072
%RSD	.47123	.44745	.37273	.15954	1.3288	.57493	3.3395	.11822
#1	.49753	76.664	.23461	4.7848	.23800	.57665	.18760	6.1339
#2	.49574	77.327	.23426	4.7754	.23176	.57786	.17971	6.1201
#3	.49290	77.158	.23296	4.7697	.23495	.58292	.19200	6.1230
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098631S Acquired: 6/25/2016 21:12:02 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47308**	**2.3488**	**.46537**	**.23713**	**.48550**	**.47406**	**.01603**
Stddev	.00180	.0057	.00826	.00130	.00209	.00072	.00324
%RSD	.38025	.24432	1.7741	.54866	.42967	.15151	20.234
#1	.47478	2.3426	.46034	.23744	.48500	.47482	.01233
#2	.47325	2.3539	.46086	.23825	.48371	.47397	.01737
#3	.47120	2.3498	.47490	.23571	.48779	.47340	.01839
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9509.0**	**58216.**	**3011.6**
Stddev	7.0	203.	15.7
%RSD	.07407	.34863	.52135
#1	9501.0	58067.	3028.6
#2	9511.6	58447.	2997.6
#3	9514.3	58134.	3008.7

Approved: June 27, 2016

Sample Name: L1606098633SD Acquired: 6/25/2016 21:15:41 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-06

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19322**	**4.9340**	**.19821**	**1.0094**	**.52302**	**.02438**	**139.13**	**.02459**
Stddev	.00378	.0145	.00351	.0014	.00226	.00009	.67	.00034
%RSD	1.9571	.29336	1.7708	.13572	.43296	.36107	.48341	1.3705
#1	.18918	4.9313	.20201	1.0102	.52283	.02442	139.33	.02427
#2	.19667	4.9211	.19752	1.0078	.52537	.02443	139.68	.02494
#3	.19382	4.9497	.19509	1.0102	.52086	.02428	138.38	.02455
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09514**	**.23910**	**.25336**	**2.0284**	**26.984**	**.53592**	**39.198**	**.43704**
Stddev	.00042	.00077	.00150	.0114	.096	.00060	.184	.00451
%RSD	.44105	.32091	.59054	.56072	.35526	.11184	.46916	1.0308
#1	.09541	.23859	.25475	2.0343	27.015	.53581	38.996	.43232
#2	.09535	.23873	.25356	2.0357	27.059	.53539	39.242	.44129
#3	.09465	.23999	.25178	2.0153	26.876	.53657	39.356	.43751
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50931**	**78.441**	**.23759**	**4.8949**	**.24046**	**.58794**	**.19015**	**6.2655**
Stddev	.00186	.336	.00142	.0178	.00324	.00596	.00110	.0092
%RSD	.36520	.42898	.59954	.36432	1.3470	1.0133	.57938	.14668
#1	.51115	78.549	.23885	4.9125	.24157	.58732	.19136	6.2719
#2	.50933	78.710	.23604	4.8953	.23682	.59419	.18990	6.2695
#3	.50743	78.063	.23787	4.8768	.24301	.58232	.18920	6.2549
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098633SD Acquired: 6/25/2016 21:15:41 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-06

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48269**	**2.3976**	**.47749**	**.23961**	**.50144**	**.48370**	**.00655**
Stddev	.00278	.0083	.00742	.00256	.00242	.00154	.00139
%RSD	.57608	.34786	1.5549	1.0679	.48318	.31822	21.270
#1	.48505	2.3997	.47422	.24207	.50191	.48531	.00572
#2	.48340	2.4047	.47226	.23981	.49882	.48355	.00816
#3	.47963	2.3884	.48599	.23696	.50360	.48224	.00577
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9541.3**	**57975.**	**3017.9**
Stddev	6.7	58.	25.7
%RSD	.06977	.09979	.85022
#1	9537.4	57946.	3002.1
#2	9549.0	58042.	3004.1
#3	9537.6	57938.	3047.5

Approved: June 27, 2016

Sample Name: L1606098635S Acquired: 6/25/2016 21:19:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-07

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.19277	4.9885	.19579	1.3512	.50932	.02402	246.59	.02462
Stddev	.00213	.0073	.00137	.0073	.00284	.00001	1.33	.00040
%RSD	1.1047	.14711	.70076	.54069	.55753	.03190	.53842	1.6295
#1	.19260	4.9802	.19687	1.3428	.50650	.02402	245.12	.02488
#2	.19498	4.9915	.19425	1.3565	.50929	.02401	246.97	.02416
#3	.19074	4.9939	.19626	1.3542	.51218	.02403	247.69	.02483
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09311	.23510	.26651	1.9710	30.025	.51865	72.396	.67936
Stddev	.00111	.00066	.00283	.0124	.152	.00282	.624	.00361
%RSD	1.1973	.28134	1.0623	.63080	.50639	.54413	.86161	.53080
#1	.09375	.23474	.26754	1.9611	29.892	.51908	71.843	.67568
#2	.09182	.23470	.26331	1.9669	29.993	.51564	72.272	.67951
#3	.09375	.23586	.26869	1.9849	30.191	.52123	73.072	.68289
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.49951	91.017	.23206	4.8553	.23760	.58301	.18517	6.5213
Stddev	.00475	.491	.00241	.0437	.00246	.00711	.00329	.0728
%RSD	.95179	.53976	1.0367	.89938	1.0354	1.2190	1.7758	1.1160
#1	.50253	90.509	.23402	4.8765	.23815	.58650	.18639	6.5577
#2	.49403	91.051	.22938	4.8051	.23491	.57483	.18144	6.4375
#3	.50197	91.490	.23278	4.8843	.23974	.58769	.18767	6.5687
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098635S Acquired: 6/25/2016 21:19:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-07

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47340**	**2.1822**	**.45682**	**.23569**	**.48957**	**.47212**	**.00440**
Stddev	.00586	.0116	.00441	.00472	.00317	.00386	.00078
%RSD	1.2388	.53009	.96525	2.0030	.64754	.81664	17.679
#1	.47626	2.1692	.45177	.23827	.48595	.47646	.00457
#2	.46666	2.1856	.45986	.23024	.49090	.46909	.00508
#3	.47729	2.1916	.45885	.23856	.49186	.47080	.00355
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9474.0**	**57681.**	**2990.6**
Stddev	86.2	193.	10.0
%RSD	.90991	.33483	.33371
#1	9421.7	57902.	3000.6
#2	9573.5	57545.	2990.5
#3	9426.8	57595.	2980.6

Approved: June 27, 2016

Sample Name: L1606098637SD Acquired: 6/25/2016 21:22:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-08

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.18854**	**4.9112**	**.19745**	**1.3462**	**.50578**	**.02318**	**246.84**	**.02457**
Stddev	.00237	.0122	.00114	.0020	.00229	.00008	1.01	.00037
%RSD	1.2581	.24857	.57516	.14562	.45291	.33396	.40736	1.5133
#1	.18824	4.9085	.19618	1.3456	.50836	.02311	247.86	.02498
#2	.18633	4.9245	.19780	1.3447	.50496	.02326	245.85	.02445
#3	.19105	4.9006	.19837	1.3484	.50401	.02315	246.80	.02427
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09368**	**.23124**	**.24645**	**1.9652**	**29.866**	**.51243**	**71.961**	**.67062**
Stddev	.00032	.00174	.00162	.0103	.162	.00260	.245	.00120
%RSD	.34157	.75435	.65887	.52496	.54204	.50740	.33982	.17871
#1	.09405	.23044	.24793	1.9540	30.010	.51110	72.114	.67080
#2	.09349	.23325	.24471	1.9744	29.691	.51543	71.679	.66935
#3	.09350	.23005	.24671	1.9671	29.897	.51077	72.089	.67172
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50362**	**90.574**	**.23429**	**4.8998**	**.23803**	**.58838**	**.19013**	**6.6122**
Stddev	.00184	.311	.00137	.0310	.00392	.00166	.00571	.0196
%RSD	.36548	.34377	.58644	.63360	1.6481	.28131	3.0018	.29638
#1	.50535	90.891	.23522	4.9314	.24008	.58995	.18521	6.6237
#2	.50169	90.269	.23272	4.8693	.24051	.58666	.19639	6.5896
#3	.50382	90.562	.23494	4.8987	.23351	.58854	.18878	6.6233
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098637SD Acquired: 6/25/2016 21:22:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573544-08

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47518**	**2.1930**	**.45556**	**.23535**	**.47970**	**.47047**	**.00421**
Stddev	.00224	.0050	.00619	.00164	.00083	.00231	.00126
%RSD	.47229	.22774	1.3595	.69822	.17220	.49043	29.868
#1	.47661	2.1987	.46015	.23423	.48048	.47206	.00301
#2	.47259	2.1891	.45801	.23458	.47883	.46782	.00552
#3	.47634	2.1914	.44851	.23723	.47980	.47152	.00409
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9517.7**	**58425.**	**3034.0**
Stddev	30.6	371.	4.3
%RSD	.32191	.63434	.14141
#1	9523.8	58332.	3029.9
#2	9544.8	58110.	3038.5
#3	9484.5	58834.	3033.6

Approved: June 27, 2016

Sample Name: L1606098639 Acquired: 6/25/2016 21:26:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00004	.00422	.00040	.00975	.00108	.00009	.14419	-.00005
Stddev	.00042	.00758	.00073	.00126	.00087	.00008	.01744	.00009
%RSD	1118.0	179.89	181.74	12.897	80.476	90.859	12.096	178.96
#1	.00028	.01060	-.00028	.01106	.00209	.00013	.16280	-.00014
#2	.00012	-.00417	.00116	.00964	.00057	-.00000	.14155	-.00004
#3	-.00051	.00622	.00032	.00856	.00059	.00013	.12822	.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00044	.00035	.00305	.04079	-.00706	.00203	-.02733	-.00024
Stddev	.00042	.00030	.00090	.02484	.04162	.00092	.03833	.00264
%RSD	93.882	85.053	29.494	60.907	589.18	45.533	140.24	1117.6
#1	.00013	.00037	.00205	.05000	-.05467	.00295	-.00835	-.00277
#2	.00091	.00004	.00332	.05971	.01106	.00204	-.00220	-.00044
#3	.00028	.00063	.00378	.01266	.02242	.00110	-.07145	.00250
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00046	.23384	.00002	.00301	.00223	-.00336	.00424	.10834
Stddev	.00025	.02078	.00057	.00488	.00185	.00118	.00213	.00216
%RSD	54.203	8.8874	3135.0	161.94	83.035	35.176	50.129	1.9925
#1	.00023	.25768	.00001	-.00120	.00258	-.00441	.00208	.11078
#2	.00073	.22428	.00059	.00188	.00389	-.00208	.00633	.10755
#3	.00043	.21955	-.00055	.00835	.00023	-.00360	.00432	.10668
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098639 Acquired: 6/25/2016 21:26:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00036	.00188	.00201	-.00117	.00072	.00941	.00130
Stddev	.00073	.00051	.00315	.00081	.00019	.00001	.00060
%RSD	203.49	27.359	156.60	69.426	26.854	.12705	46.035
#1	.00061	.00146	.00564	-.00068	.00081	.00940	.00116
#2	.00093	.00173	.00038	-.00211	.00085	.00941	.00078
#3	-.00047	.00245	.00001	-.00072	.00050	.00942	.00196
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9979.7	61621.	3102.9
Stddev	10.2	118.	6.8
%RSD	.10173	.19116	.21926
#1	9983.1	61738.	3101.0
#2	9987.7	61621.	3110.5
#3	9968.3	61503.	3097.3

Approved: June 27, 2016

Sample Name: L1606098641 Acquired: 6/25/2016 21:30:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00155	.00228	-.00074	.00681	.00240	.00007	2.3285	.00015
Stddev	.00195	.00437	.00132	.00259	.00010	.00002	.0225	.00014
%RSD	125.93	191.41	178.17	38.016	4.0011	29.007	.96410	94.650
#1	.00016	.00624	-.00146	.00963	.00251	.00005	2.3542	.00003
#2	.00071	.00302	-.00156	.00453	.00235	.00009	2.3184	.00011
#3	.00378	-.00241	.00078	.00628	.00234	.00008	2.3128	.00030
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00006	.00014	.00163	.03413	.24665	.00092	.81266	-.00079
Stddev	.00068	.00062	.00075	.01156	.02854	.00423	.01745	.00318
%RSD	1106.1	456.64	45.829	33.860	11.573	458.16	2.1475	404.35
#1	-.00033	.00085	.00092	.03095	.24915	.00580	.81529	.00231
#2	.00085	-.00023	.00241	.02450	.27386	-.00166	.79404	-.00404
#3	-.00034	-.00022	.00156	.04694	.21693	-.00138	.82864	-.00063
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	1.5234	-.00034	.00759	.00193	-.00110	-.00056	1.9225
Stddev	.00028	.0181	.00053	.00504	.00328	.00190	.00524	.0055
%RSD	316.47	1.1895	155.07	66.429	170.02	173.25	938.70	.28386
#1	-.00021	1.5412	-.00096	.00263	.00004	-.00329	-.00638	1.9218
#2	.00035	1.5240	.00002	.00743	.00571	-.00014	.00092	1.9283
#3	.00013	1.5050	-.00009	.01271	.00003	.00014	.00379	1.9175
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606098641 Acquired: 6/25/2016 21:30:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00003**	**.02979**	**.00925**	**-.00425**	**.00079**	**.00575**	**.00161**
Stddev	.00070	.00033	.00260	.00161	.00118	.00006	.00128
%RSD	2613.1	1.1094	28.082	37.920	148.25	1.0430	79.495
#1	-.00014	.02957	.00917	-.00541	.00147	.00579	.00219
#2	.00072	.03017	.01188	-.00492	.00148	.00578	.00014
#3	-.00067	.02963	.00669	-.00241	-.00057	.00568	.00250
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9984.0**	**61505.**	**3103.5**
Stddev	25.4	97.	8.6
%RSD	.25425	.15693	.27670
#1	10013.	61422.	3093.7
#2	9966.4	61611.	3106.9
#3	9972.6	61483.	3109.8

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 21:34:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.40382	10.228	.39903	.49252	1.0184	.05057	10.132
Stddev	.00317	.045	.00339	.00303	.0029	.00010	.084
%RSD	.78510	.44126	.84879	.61522	.28786	.20129	.82686
#1	.40329	10.193	.40176	.48985	1.0165	.05045	10.103
#2	.40723	10.279	.39524	.49189	1.0170	.05064	10.066
#3	.40095	10.211	.40009	.49581	1.0218	.05061	10.226
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05109	.20032	.50425	.51389	3.9993	49.456	1.0271
Stddev	.00034	.00030	.00243	.00134	.0288	.230	.0087
%RSD	.66912	.15054	.48104	.26170	.72080	.46581	.84751
#1	.05096	.20027	.50150	.51248	3.9697	49.237	1.0193
#2	.05148	.20064	.50519	.51515	4.0273	49.435	1.0365
#3	.05084	.20004	.50607	.51404	4.0011	49.697	1.0255
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.5965	.50354	1.0171	50.205	.50852	9.9839	.50630
Stddev	.0605	.00125	.0104	.188	.00105	.0190	.00443
%RSD	.63000	.24843	1.0188	.37433	.20655	.18977	.87552
#1	9.5498	.50228	1.0149	50.026	.50872	9.9714	.50177
#2	9.5748	.50478	1.0283	50.189	.50945	10.006	.51063
#3	9.6648	.50354	1.0080	50.400	.50738	9.9747	.50650
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 21:34:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2057**	**.39535**	**5.1622**	**1.0068**	**1.0152**	**1.0163**	**.51868**
Stddev	.0009	.00724	.0066	.0035	.0056	.0047	.00189
%RSD	.07645	1.8317	.12822	.34938	.55282	.46463	.36343
#1	1.2067	.40045	5.1563	1.0056	1.0088	1.0109	.51803
#2	1.2055	.39854	5.1610	1.0107	1.0174	1.0194	.51721
#3	1.2048	.38706	5.1694	1.0040	1.0193	1.0187	.52081
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0297**	**1.0108**	F **2.9654**
Stddev	.0010	.0013	.0124
%RSD	.09618	.12981	.41944
#1	1.0297	1.0109	2.9530
#2	1.0307	1.0120	2.9654
#3	1.0287	1.0094	2.9779
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9765.3**	**59222.**	**3028.4**
Stddev	12.6	103.	19.9
%RSD	.12950	.17379	.65564
#1	9777.2	59331.	3048.9
#2	9752.0	59127.	3027.1
#3	9766.5	59207.	3009.2

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 21:38:07 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00048**	**-.00056**	**.00159**	**.00238**	**.00173**	**.00014**	**-.05704**	**-.00011**
Stddev	.00089	.00372	.00374	.00230	.00036	.00002	.05178	.00017
%RSD	185.89	660.56	234.30	96.559	20.708	14.400	90.783	156.11
#1	-.00133	.00080	.00118	.00024	.00195	.00012	-.11682	-.00027
#2	-.00055	-.00477	-.00192	.00481	.00132	.00016	-.02846	-.00014
#3	.00045	.00228	.00552	.00209	.00192	.00015	-.02584	.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00017**	**.00048**	**.00001**	**.01554**	**-.08094**	**.00031**	**-.00960**	**.00059**
Stddev	.00009	.00037	.00173	.01679	.11260	.00565	.05873	.00171
%RSD	56.610	77.335	20230.	108.07	139.11	1853.2	611.58	288.38
#1	.00024	.00066	.00194	-.00104	.01136	.00158	-.02063	.00007
#2	.00020	.00072	-.00054	.03254	-.20639	-.00588	.05386	.00251
#3	.00006	.00005	-.00138	.01511	-.04780	.00521	-.06203	-.00080
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00336**	**.02885**	**-.00050**	**.00712**	**.00213**	**.00602**	**-.00054**	**.04214**
Stddev	.00024	.02974	.00074	.00312	.00180	.00158	.00069	.00013
%RSD	7.0582	103.10	148.29	43.907	84.214	26.247	127.12	.29697
#1	.00323	-.00548	.00008	.00759	.00359	.00782	.00008	.04201
#2	.00321	.04496	-.00024	.00998	.00013	.00485	-.00042	.04226
#3	.00363	.04705	-.00133	.00378	.00268	.00540	-.00129	.04215
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 21:38:07 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00047	.00065	-.00177	-.00021	.00113	.00005	.00375
Stddev	.00100	.00044	.00221	.00418	.00029	.00020	.00087
%RSD	212.41	68.047	125.42	1958.4	25.944	435.57	23.135
#1	.00160	.00080	.00075	-.00080	.00139	-.00005	.00437
#2	-.00031	.00015	-.00341	-.00407	.00081	.00028	.00276
#3	.00012	.00100	-.00264	.00422	.00119	-.00009	.00413
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9540.4	58573.	2934.4
Stddev	6.2	75.	17.1
%RSD	.06498	.12729	.58178
#1	9533.2	58646.	2919.8
#2	9543.3	58575.	2953.2
#3	9544.6	58497.	2930.3

Approved: June 27, 2016

Sample Name: PBW 9A Acquired: 6/25/2016 21:42:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**.00614**	**.00175**	**.00056**	**.00231**	**.00012**	**-.04646**	**-.00007**
Stddev	.00171	.00204	.00393	.00252	.00071	.00020	.04124	.00026
%RSD	1283.4	33.141	224.26	449.74	30.683	160.24	88.765	365.50
#1	.00182	.00611	.00626	-.00197	.00237	.00018	-.00258	-.00036
#2	-.00137	.00820	-.00088	.00057	.00157	.00029	-.08442	.00014
#3	-.00085	.00413	-.00013	.00307	.00298	-.00010	-.05240	.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00037**	**.00010**	**.00080**	**.02407**	**-.12665**	**.00141**	**-.00997**	**-.00231**
Stddev	.00033	.00034	.00081	.01192	.12266	.00378	.02213	.00072
%RSD	90.436	329.64	101.91	49.545	96.850	267.21	222.01	31.042
#1	-.00000	-.00017	-.00014	.03113	-.25310	.00322	.01487	-.00193
#2	-.00045	.00048	.00118	.03078	-.11869	.00395	-.02759	-.00185
#3	-.00065	-.00000	.00135	.01030	-.00816	-.00293	-.01718	-.00313
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00037**	**.03613**	**.00006**	**.00672**	**.00326**	**-.00431**	**-.00184**	**.00566**
Stddev	.00007	.03275	.00077	.00152	.00253	.00485	.00591	.00059
%RSD	18.235	90.627	1212.2	22.673	77.664	112.59	322.28	10.478
#1	.00035	.07347	.00060	.00612	.00546	-.00044	-.00125	.00618
#2	.00032	.02263	-.00082	.00846	.00049	-.00975	.00376	.00501
#3	.00045	.01230	.00041	.00559	.00385	-.00274	-.00802	.00578
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: PBW 9A Acquired: 6/25/2016 21:42:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00063**	**.00033**	**-.00143**	**-.00317**	**.00125**	**.00161**	**.00218**
Stddev	.00053	.00015	.00130	.00329	.00039	.00011	.00081
%RSD	84.270	46.594	90.758	103.97	30.982	7.0758	37.192
#1	.00094	.00050	-.00075	-.00532	.00080	.00148	.00195
#2	.00002	.00020	-.00062	-.00481	.00145	.00171	.00150
#3	.00092	.00028	-.00293	.00062	.00150	.00162	.00308
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9585.8**	**59115.**	**2963.7**
Stddev	26.3	98.	6.0
%RSD	.27487	.16506	.20388
#1	9574.9	59148.	2962.9
#2	9615.8	59005.	2958.2
#3	9566.6	59192.	2970.2

Approved: June 27, 2016

Sample Name: LCSW 9A Acquired: 6/25/2016 21:45:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19278**	**4.8645**	**.19357**	**.92620**	**.49567**	**.02388**	**4.9744**	**.02438**
Stddev	.00059	.0150	.00172	.00296	.00123	.00001	.0131	.00027
%RSD	.30529	.30879	.88618	.31973	.24905	.05168	.26315	1.0959
#1	.19266	4.8773	.19194	.92961	.49521	.02387	4.9745	.02469
#2	.19342	4.8480	.19536	.92439	.49707	.02389	4.9613	.02424
#3	.19226	4.8682	.19339	.92458	.49473	.02389	4.9875	.02421
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09756**	**.24011**	**.24908**	**1.9287**	**23.845**	**.50347**	**4.7458**	**.24452**
Stddev	.00018	.00063	.00057	.0379	.076	.00304	.1100	.00304
%RSD	.18802	.26044	.22811	1.9678	.32025	.60437	2.3184	1.2446
#1	.09771	.24070	.24970	1.9204	23.902	.50653	4.7834	.24200
#2	.09761	.23946	.24858	1.9701	23.875	.50343	4.6218	.24790
#3	.09735	.24018	.24895	1.8956	23.758	.50044	4.8320	.24367
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50361**	**24.550**	**.24652**	**4.6792**	**.24585**	**.57643**	**.18945**	**2.4007**
Stddev	.00079	.045	.00091	.0152	.00374	.00321	.00289	.0038
%RSD	.15600	.18503	.37091	.32489	1.5220	.55613	1.5253	.15739
#1	.50355	24.561	.24748	4.6824	.24789	.57981	.19277	2.4049
#2	.50442	24.588	.24565	4.6926	.24153	.57604	.18748	2.3993
#3	.50285	24.500	.24642	4.6627	.24812	.57343	.18811	2.3978
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LCSW 9A Acquired: 6/25/2016 21:45:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48346**	**.50078**	**.49876**	**.24644**	**.49280**	**.48316**	**.01652**
Stddev	.00157	.00182	.00430	.00290	.00093	.00040	.00416
%RSD	.32431	.36335	.86149	1.1750	.18883	.08190	25.157
#1	.48465	.50112	.49719	.24479	.49339	.48270	.01275
#2	.48404	.50241	.49547	.24978	.49173	.48334	.01583
#3	.48168	.49882	.50362	.24474	.49328	.48342	.02097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9876.1**	**60275.**	**3070.8**
Stddev	6.3	180.	12.0
%RSD	.06355	.29822	.39033
#1	9874.5	60221.	3071.7
#2	9883.0	60476.	3058.4
#3	9870.8	60129.	3082.3

Approved: June 27, 2016

Sample Name: F BLANK Acquired: 6/25/2016 21:49:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00055	-.00126	-.00015	.00486	.00128	.00012	-.00459	.00000
Stddev	.00056	.00100	.00110	.00023	.00095	.00001	.02969	.00020
%RSD	100.53	79.131	736.67	4.6731	74.371	6.7718	646.26	7554.0
#1	.00064	-.00237	-.00130	.00509	.00028	.00012	.02032	.00019
#2	.00106	-.00043	-.00004	.00484	.00217	.00011	-.03745	-.00020
#3	-.00004	-.00099	.00090	.00463	.00139	.00012	.00335	.00002
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00037	.00026	.00088	.03134	-.11502	-.00026	.03535	-.00044
Stddev	.00018	.00121	.00068	.00659	.05233	.00265	.10514	.00096
%RSD	49.613	468.36	76.871	21.019	45.492	1017.6	297.45	218.87
#1	.00058	.00075	.00159	.02374	-.13830	.00130	-.07629	.00060
#2	.00024	.00114	.00082	.03485	-.15168	.00124	.13249	-.00129
#3	.00028	-.00112	.00024	.03543	-.05510	-.00331	.04985	-.00062
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00056	.07370	.00132	.00313	.00117	-.00091	.00077	-.00309
Stddev	.00026	.00588	.00035	.00217	.00118	.00090	.00482	.00173
%RSD	45.817	7.9801	26.636	69.209	100.20	98.397	629.48	56.034
#1	.00085	.06695	.00170	.00515	.00095	-.00029	-.00041	-.00378
#2	.00042	.07645	.00125	.00084	.00245	-.00194	-.00336	-.00438
#3	.00040	.07770	.00101	.00340	.00013	-.00050	.00606	-.00112
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: F BLANK Acquired: 6/25/2016 21:49:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00033**	**.00302**	**-.00048**	**-.00333**	**.00002**	**.00245**	**.00132**
Stddev	.00077	.00023	.00928	.00236	.00016	.00009	.00057
%RSD	230.12	7.5634	1914.7	70.982	654.43	3.7966	42.974
#1	.00023	.00325	-.01051	-.00146	-.00016	.00251	.00071
#2	.00115	.00279	.00127	-.00253	.00008	.00234	.00183
#3	-.00038	.00303	.00779	-.00598	.00016	.00250	.00143
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9948.0**	**61128.**	**3082.2**
Stddev	8.3	35.	19.2
%RSD	.08376	.05747	.62310
#1	9943.7	61136.	3060.8
#2	9942.7	61089.	3097.9
#3	9957.6	61158.	3087.8

Approved: June 27, 2016

Sample Name: L1606078709 Acquired: 6/25/2016 21:53:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00106	.32067	.00199	.03822	.08666	.00012	52.298	.00025
Stddev	.00033	.00225	.00178	.00083	.00065	.00013	.384	.00005
%RSD	31.425	.70249	89.821	2.1841	.74667	104.78	.73398	21.021
#1	.00100	.32316	.00023	.03790	.08700	-.00002	52.414	.00024
#2	.00077	.32010	.00380	.03760	.08707	.00018	51.870	.00021
#3	.00143	.31876	.00193	.03917	.08592	.00021	52.611	.00031
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00014	.00232	.00233	.31407	.84091	.00625	16.422	.01281
Stddev	.00025	.00123	.00162	.01381	.09733	.00295	.087	.00086
%RSD	177.99	52.851	69.798	4.3981	11.574	47.203	.53048	6.7172
#1	-.00001	.00331	.00164	.32378	.73706	.00531	16.422	.01228
#2	.00000	.00271	.00116	.32018	.93004	.00389	16.510	.01236
#3	.00043	.00095	.00418	.29826	.85562	.00956	16.336	.01381
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00023	3.5331	.00030	.18004	.00409	-.00180	.00164	8.5479
Stddev	.00032	.0286	.00045	.00328	.00158	.00258	.00446	.0078
%RSD	137.02	.80985	147.61	1.8192	38.579	143.43	271.62	.09172
#1	-.00011	3.5554	-.00012	.17874	.00498	-.00359	.00673	8.5558
#2	.00052	3.5008	.00077	.17762	.00503	-.00297	-.00020	8.5477
#3	.00028	3.5432	.00026	.18377	.00227	.00116	-.00161	8.5401
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078709 Acquired: 6/25/2016 21:53:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00032	.13816	-.00281	-.00107	.00371	.00875	.00154
Stddev	.00042	.00078	.00681	.00141	.00081	.00007	.00044
%RSD	132.89	.56285	242.62	131.61	21.926	.80395	28.630
#1	-.00012	.13842	.00500	-.00130	.00416	.00881	.00162
#2	.00035	.13728	-.00586	-.00235	.00277	.00878	.00193
#3	.00072	.13877	-.00757	.00044	.00421	.00867	.00106
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9876.5	60888.	3069.7
Stddev	11.3	192.	3.6
%RSD	.11437	.31559	.11816
#1	9889.0	60702.	3066.7
#2	9866.9	60874.	3073.8
#3	9873.6	61086.	3068.7

Approved: June 27, 2016

Sample Name: L1606078711S Acquired: 6/25/2016 21:57:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19226**	**5.2769**	**.19700**	**.97637**	**.57407**	**.02377**	**55.412**	**.02464**
Stddev	.00232	.0172	.00210	.00695	.00123	.00006	.275	.00018
%RSD	1.2075	.32678	1.0663	.71175	.21357	.23547	.49636	.71677
#1	.19492	5.2928	.19484	.98432	.57271	.02383	55.376	.02446
#2	.19066	5.2586	.19711	.97142	.57439	.02373	55.156	.02465
#3	.19119	5.2792	.19903	.97337	.57510	.02375	55.703	.02481
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09590**	**.23900**	**.24525**	**2.2503**	**24.873**	**.51320**	**20.375**	**.25171**
Stddev	.00047	.00136	.00161	.0156	.173	.01009	.023	.00267
%RSD	.48644	.56990	.65491	.69491	.69517	1.9665	.11329	1.0594
#1	.09633	.24013	.24639	2.2591	25.016	.52369	20.389	.25241
#2	.09595	.23749	.24595	2.2322	24.681	.50356	20.349	.25396
#3	.09541	.23938	.24342	2.2595	24.921	.51234	20.388	.24877
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50427**	**27.841**	**.24087**	**4.9542**	**.24449**	**.58622**	**.18341**	**11.076**
Stddev	.00227	.102	.00131	.0062	.00416	.00262	.00666	.019
%RSD	.45102	.36806	.54487	.12407	1.6996	.44696	3.6299	.17466
#1	.50494	27.874	.24041	4.9608	.24621	.58442	.18291	11.063
#2	.50614	27.726	.24235	4.9486	.23975	.58502	.19031	11.099
#3	.50174	27.923	.23985	4.9531	.24750	.58923	.17702	11.068
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078711S Acquired: 6/25/2016 21:57:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48537**	**.63044**	**.49547**	**.24660**	**.49048**	**.48728**	**.00709**
Stddev	.00041	.00808	.00649	.00259	.00192	.00292	.00024
%RSD	.08494	1.2823	1.3100	1.0500	.39154	.59943	3.4352
#1	.48528	.63362	.48907	.24492	.49247	.49009	.00732
#2	.48501	.63645	.50205	.24958	.48864	.48426	.00711
#3	.48582	.62125	.49531	.24529	.49034	.48750	.00684
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9767.2**	**59939.**	**3081.9**
Stddev	8.4	151.	2.2
%RSD	.08617	.25130	.07156
#1	9776.8	59826.	3083.9
#2	9761.1	60110.	3082.2
#3	9763.8	59880.	3079.5

Approved: June 27, 2016

Sample Name: L1606078712SD Acquired: 6/25/2016 22:01:06 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-06

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19422**	**6.5547**	**.19616**	**.97909**	**.58447**	**.02412**	**57.272**	**.02449**
Stddev	.00043	.0174	.00206	.00320	.00356	.00008	.255	.00025
%RSD	.22056	.26500	1.0502	.32643	.60985	.31860	.44461	1.0216
#1	.19468	6.5722	.19401	.98024	.58211	.02413	57.018	.02471
#2	.19383	6.5546	.19811	.97548	.58857	.02404	57.527	.02422
#3	.19415	6.5375	.19636	.98156	.58273	.02419	57.272	.02454
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09680**	**.24221**	**.24688**	**3.2519**	**25.085**	**.51256**	**21.272**	**.29044**
Stddev	.00024	.00084	.00160	.0305	.123	.00182	.123	.00255
%RSD	.25204	.34673	.64632	.93681	.48984	.35468	.57596	.87927
#1	.09682	.24178	.24538	3.2175	24.965	.51125	21.131	.28769
#2	.09704	.24318	.24856	3.2756	25.081	.51179	21.342	.29274
#3	.09655	.24168	.24672	3.2624	25.210	.51464	21.344	.29089
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50646**	**28.018**	**.24386**	**4.9818**	**.24358**	**.58240**	**.18769**	**13.764**
Stddev	.00076	.167	.00147	.0053	.00228	.00382	.00425	.008
%RSD	.15029	.59670	.60238	.10534	.93711	.65545	2.2629	.06131
#1	.50691	27.825	.24251	4.9763	.24111	.57903	.18298	13.756
#2	.50690	28.112	.24543	4.9823	.24561	.58163	.18886	13.772
#3	.50558	28.117	.24364	4.9868	.24402	.58655	.19123	13.765
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078712SD Acquired: 6/25/2016 22:01:06 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573102-06

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48752**	**.63979**	**.50344**	**.24632**	**.49636**	**.48329**	**.00682**
Stddev	.00123	.00383	.00416	.00077	.00180	.00044	.00097
%RSD	.25232	.59818	.82653	.31336	.36344	.09176	14.263
#1	.48785	.63646	.50591	.24591	.49538	.48373	.00713
#2	.48616	.64397	.49864	.24584	.49525	.48329	.00761
#3	.48856	.63895	.50578	.24721	.49844	.48284	.00573
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9745.6**	**59867.**	**3081.8**
Stddev	9.9	91.	26.0
%RSD	.10194	.15198	.84468
#1	9734.5	59851.	3110.8
#2	9753.6	59965.	3060.5
#3	9748.6	59785.	3074.0

Approved: June 27, 2016

Sample Name: L1606078701 Acquired: 6/25/2016 22:04:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00206	-.00089	-.00017	.00563	.00151	.00011	.08615	.00015
Stddev	.00234	.00167	.00131	.00160	.00060	.00001	.02054	.00021
%RSD	113.42	187.14	751.31	28.470	39.897	9.3705	23.846	140.35
#1	.00316	-.00269	.00118	.00529	.00196	.00012	.07852	-.00008
#2	-.00062	-.00061	-.00144	.00737	.00083	.00010	.07052	.00033
#3	.00364	.00062	-.00027	.00422	.00173	.00012	.10942	.00019
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00013	.00093	.00322	.06115	.01246	.00241	.02800	.00003
Stddev	.00003	.00076	.00095	.00519	.08806	.00216	.01589	.00083
%RSD	20.822	82.163	29.586	8.4819	706.92	89.891	56.731	2625.4
#1	.00011	.00029	.00408	.06573	-.03307	.00026	.04634	-.00085
#2	.00017	.00177	.00219	.05552	.11396	.00458	.01831	.00015
#3	.00013	.00071	.00338	.06220	-.04352	.00238	.01936	.00079
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00040	.44862	.00009	.00157	.00319	-.00012	.00644	.10535
Stddev	.00015	.01169	.00041	.00200	.00252	.00436	.00247	.00160
%RSD	36.378	2.6064	429.23	127.80	78.785	3632.8	38.299	1.5199
#1	.00041	.45408	.00056	-.00069	.00238	-.00321	.00918	.10450
#2	.00025	.45659	-.00021	.00226	.00118	.00487	.00438	.10720
#3	.00054	.43520	-.00006	.00313	.00602	-.00202	.00577	.10436
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078701 Acquired: 6/25/2016 22:04:44 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00114	.00134	.00464	-.00236	.00126	.00852	.00204
Stddev	.00023	.00031	.01117	.00170	.00022	.00013	.00133
%RSD	19.939	23.344	240.52	72.025	17.171	1.4749	64.993
#1	.00100	.00168	-.00728	-.00407	.00147	.00841	.00183
#2	.00140	.00106	.01487	-.00232	.00127	.00866	.00083
#3	.00102	.00128	.00634	-.00068	.00104	.00851	.00346
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9978.6	61725.	3108.7
Stddev	25.8	52.	12.8
%RSD	.25894	.08436	.41157
#1	9949.5	61751.	3109.2
#2	9998.7	61759.	3121.3
#3	9987.8	61665.	3095.7

Approved: June 27, 2016

Sample Name: L1606078701PS Acquired: 6/25/2016 22:08:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573561-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19562**	**4.8670**	**.19417**	**.93274**	**.49617**	**.02396**	**5.0466**	**.02449**
Stddev	.00101	.0014	.00206	.00077	.00126	.00015	.0443	.00024
%RSD	.51610	.02767	1.0615	.08264	.25431	.60837	.87718	.96002
#1	.19561	4.8655	.19531	.93216	.49577	.02379	5.0119	.02471
#2	.19461	4.8672	.19179	.93245	.49515	.02407	5.0964	.02451
#3	.19663	4.8682	.19541	.93362	.49758	.02401	5.0313	.02424
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09815**	**.24178**	**.25186**	**1.9475**	**23.902**	**.50912**	**4.6690**	**.24615**
Stddev	.00038	.00032	.00085	.0476	.079	.00280	.1745	.00176
%RSD	.38848	.13325	.33570	2.4417	.33116	.54926	3.7379	.71493
#1	.09847	.24215	.25095	2.0020	23.984	.51180	4.5199	.24606
#2	.09825	.24158	.25263	1.9144	23.825	.50622	4.6262	.24443
#3	.09773	.24160	.25200	1.9261	23.897	.50934	4.8609	.24795
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.50547**	**24.979**	**.24871**	**4.7191**	**.25003**	**.58513**	**.18331**	**2.5397**
Stddev	.00148	.047	.00145	.0045	.00153	.00133	.00963	.0085
%RSD	.29370	.18695	.58415	.09482	.61173	.22744	5.2541	.33396
#1	.50679	24.962	.24995	4.7237	.24937	.58423	.17385	2.5470
#2	.50576	24.943	.24906	4.7186	.24895	.58665	.18298	2.5417
#3	.50387	25.032	.24711	4.7148	.25178	.58449	.19310	2.5304
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078701PS Acquired: 6/25/2016 22:08:42 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573561-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.48806	.50437	.49939	.24983	.48939	.48880	.00250
Stddev	.00273	.00349	.00591	.00248	.00296	.00032	.00045
%RSD	.55970	.69122	1.1836	.99157	.60527	.06447	18.096
#1	.48918	.50154	.50066	.25202	.49256	.48872	.00242
#2	.49005	.50331	.49295	.25032	.48893	.48914	.00298
#3	.48494	.50827	.50457	.24714	.48669	.48853	.00209
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9876.1	60662.	3087.6
Stddev	12.0	38.	4.3
%RSD	.12163	.06253	.13975
#1	9869.1	60619.	3083.6
#2	9869.3	60677.	3092.2
#3	9890.0	60690.	3087.1

Approved: June 27, 2016

Sample Name: L1606078701SDL Acquired: 6/25/2016 22:12:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573561-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00226	-.00962	.00257	.00372	.00008	.00010	-.05467	-.00005
Stddev	.00390	.01062	.00091	.00131	.00131	.00006	.03854	.00004
%RSD	172.35	110.37	35.549	35.155	1747.1	57.750	70.491	84.699
#1	.00675	-.01272	.00350	.00225	-.00089	.00014	-.04307	-.00000
#2	.00027	-.01835	.00254	.00475	.00157	.00012	-.09767	-.00007
#3	-.00023	.00220	.00167	.00417	-.00046	.00003	-.02326	-.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00015	.00076	.00146	.02792	-.10756	-.00176	-.01577	.00085
Stddev	.00022	.00111	.00071	.01781	.06853	.00352	.01010	.00145
%RSD	144.06	146.16	48.411	63.806	63.714	199.76	64.076	169.31
#1	.00010	.00182	.00216	.04804	-.06727	-.00153	-.01532	-.00005
#2	-.00026	-.00038	.00146	.01416	-.18668	.00163	-.02609	.00009
#3	-.00029	.00083	.00075	.02156	-.06871	-.00539	-.00589	.00252
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00035	.12025	-.00037	.00742	.00388	-.00036	.00693	.02501
Stddev	.00028	.02622	.00009	.00699	.00334	.00491	.00079	.00132
%RSD	80.313	21.803	23.237	94.157	86.143	1346.9	11.341	5.2920
#1	.00057	.14797	-.00041	.00450	.00651	.00107	.00746	.02626
#2	.00045	.09585	-.00027	.01539	.00501	-.00584	.00603	.02362
#3	.00003	.11693	-.00042	.00237	.00012	.00367	.00731	.02515
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078701SDL Acquired: 6/25/2016 22:12:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG573561-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00036	.00027	.00391	-.00239	.00091	.00198	.00117
Stddev	.00062	.00058	.00259	.00083	.00057	.00021	.00105
%RSD	170.80	217.03	66.210	34.804	62.977	10.839	89.133
#1	.00032	-.00015	.00203	-.00335	.00041	.00217	.00108
#2	-.00024	.00003	.00686	-.00188	.00153	.00175	.00018
#3	.00100	.00093	.00284	-.00194	.00079	.00202	.00227
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10481.	64804.	3221.0
Stddev	25.	98.	4.2
%RSD	.23393	.15113	.13029
#1	10501.	64916.	3225.8
#2	10489.	64737.	3219.2
#3	10454.	64758.	3217.9

Approved: June 27, 2016

Sample Name: L1606078701SDL Acquired: 6/25/2016 22:16:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573561-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00050**	**-.01671**	**.00066**	**.00174**	**.00117**	**.00007**	**-.03873**	**-.00006**
Stddev	.00055	.00595	.00077	.00373	.00047	.00003	.02526	.00023
%RSD	110.50	35.627	115.52	214.66	39.929	36.665	65.224	418.10
#1	.00087	-.01063	-.00016	.00037	.00065	.00010	-.06018	.00018
#2	-.00014	-.02253	.00136	-.00111	.00156	.00005	-.01089	-.00029
#3	.00077	-.01699	.00080	.00596	.00129	.00006	-.04513	-.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00020**	**-.00027**	**.00015**	**.01537**	**-.08473**	**-.00313**	**-.04452**	**.00037**
Stddev	.00041	.00031	.00072	.00667	.12317	.00148	.02793	.00217
%RSD	208.97	115.52	489.41	43.427	145.37	47.326	62.733	580.66
#1	.00051	-.00025	-.00010	.01465	-.11440	-.00155	-.01300	-.00212
#2	-.00027	-.00059	-.00042	.02238	.05057	-.00335	-.06618	.00176
#3	.00035	.00003	.00096	.00908	-.19035	-.00448	-.05440	.00148
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00052**	**.04068**	**-.00002**	**.00630**	**.00223**	**-.00439**	**.00789**	**.00775**
Stddev	.00014	.01560	.00098	.00605	.00105	.00119	.00333	.00034
%RSD	26.596	38.347	4828.5	96.010	47.199	27.128	42.126	4.3419
#1	-.00068	.05533	-.00110	-.00065	.00121	-.00319	.01093	.00736
#2	-.00042	.02428	.00080	.00920	.00218	-.00557	.00434	.00794
#3	-.00046	.04243	.00023	.01036	.00331	-.00443	.00841	.00794
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078701SDL Acquired: 6/25/2016 22:16:16 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG573561-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00078**	**.00049**	**.00611**	**-.00169**	**-.00019**	**.00076**	**.00104**
Stddev	.00047	.00028	.00364	.00121	.00086	.00008	.00084
%RSD	59.598	55.989	59.553	71.357	443.84	10.636	80.846
#1	-.00025	.00020	.00285	-.00078	-.00105	.00067	.00138
#2	-.00100	.00075	.00545	-.00306	-.00021	.00078	.00165
#3	-.00110	.00052	.01004	-.00123	.00067	.00082	.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10473.**	**64343.**	**3185.8**
Stddev	11.	88.	33.4
%RSD	.10369	.13736	1.0486
#1	10479.	64432.	3210.0
#2	10479.	64255.	3199.9
#3	10461.	64342.	3147.7

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 22:20:12 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.39676**	**10.011**	**.39200**	**.48774**	**1.0072**	**.04931**	**9.9627**
Stddev	.00166	.042	.00060	.00247	.0020	.00016	.0336
%RSD	.41961	.41526	.15324	.50705	.19487	.32345	.33737
#1	.39852	10.036	.39268	.48499	1.0052	.04934	9.9285
#2	.39656	10.035	.39180	.48843	1.0074	.04945	9.9639
#3	.39521	9.9633	.39153	.48979	1.0091	.04914	9.9956
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05046**	**.19748**	**.49159**	**.50549**	**3.9250**	**48.860**	**1.0117**
Stddev	.00033	.00063	.00167	.00263	.0491	.051	.0023
%RSD	.66275	.32062	.34000	.52017	1.2522	.10387	.22425
#1	.05068	.19812	.49231	.50540	3.8948	48.907	1.0140
#2	.05064	.19745	.49279	.50817	3.9817	48.806	1.0118
#3	.05008	.19686	.48968	.50291	3.8986	48.867	1.0095
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.4449**	**.49486**	**1.0080**	**49.891**	**.50156**	**9.8436**	**.50078**
Stddev	.0163	.00095	.0079	.070	.00172	.0350	.00128
%RSD	.17297	.19183	.78730	.13958	.34250	.35531	.25500
#1	9.4275	.49541	1.0170	49.848	.50286	9.8791	.50070
#2	9.4600	.49376	1.0019	49.971	.50221	9.8426	.49954
#3	9.4471	.49540	1.0052	49.853	.49961	9.8092	.50209
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 22:20:12 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1892**	**.38968**	**5.0794**	**.99180**	**1.0052**	**1.0145**	**.50829**
Stddev	.0070	.00637	.0139	.00335	.0009	.0065	.00390
%RSD	.58775	1.6349	.27365	.33789	.09319	.64413	.76634
#1	1.1970	.39094	5.0928	.99452	1.0044	1.0072	.50887
#2	1.1869	.38277	5.0804	.99283	1.0062	1.0197	.51187
#3	1.1836	.39532	5.0651	.98805	1.0048	1.0166	.50414
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0022**	**.99700**	F **2.9158**
Stddev	.0042	.00301	.0038
%RSD	.42153	.30191	.13010
#1	1.0066	.99806	2.9128
#2	1.0017	.99934	2.9201
#3	.99824	.99361	2.9145
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9930.6**	**60612.**	**3086.3**
Stddev	10.1	73.	9.8
%RSD	.10207	.12111	.31687
#1	9941.9	60647.	3097.4
#2	9922.1	60528.	3082.4
#3	9927.9	60661.	3079.1

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 22:23:49 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00100**	**-.00695**	**.00023**	**.00246**	**.00203**	**.00014**	**-.01220**	**.00014**
Stddev	.00181	.00287	.00117	.00195	.00133	.00002	.03027	.00029
%RSD	181.13	41.303	500.29	79.096	65.555	16.180	248.16	212.80
#1	-.00109	-.00504	-.00098	.00360	.00305	.00017	-.03745	.00025
#2	.00205	-.00555	.00136	.00021	.00053	.00014	.02136	-.00019
#3	.00203	-.01025	.00032	.00357	.00251	.00013	-.02050	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00014**	**-.00000**	**.00097**	**.02954**	**-.02965**	**-.00129**	**.00088**	**.00163**
Stddev	.00011	.00082	.00136	.00883	.01713	.00234	.03199	.00271
%RSD	74.595	562770.	140.78	29.900	57.784	181.57	3634.1	165.69
#1	.00008	.00083	.00213	.03726	-.04191	-.00397	.03754	.00233
#2	.00008	-.00001	.00130	.01991	-.03697	-.00018	-.01355	.00393
#3	.00027	-.00081	-.00053	.03145	-.01007	.00029	-.02135	-.00135
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00354**	**.01474**	**-.00023**	**.00359**	**.00237**	**.00219**	**.00418**	**.03627**
Stddev	.00058	.04847	.00031	.00245	.00060	.00526	.00484	.00058
%RSD	16.314	328.86	136.11	68.357	25.126	240.59	115.86	1.6055
#1	.00288	-.00030	-.00040	.00322	.00284	.00365	.00822	.03616
#2	.00384	.06894	.00013	.00134	.00170	.00656	.00551	.03575
#3	.00391	-.02443	-.00042	.00620	.00257	-.00365	-.00119	.03690
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 22:23:49 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00016**	**.00039**	**.00181**	**-.00370**	**.00020**	**.00004**	**.00435**
Stddev	.00110	.00025	.00867	.00085	.00051	.00012	.00191
%RSD	689.85	64.301	479.78	22.978	253.56	271.57	43.946
#1	-.00106	.00012	.00915	-.00439	-.00031	.00014	.00375
#2	.00108	.00042	.00404	-.00397	.00020	.00008	.00281
#3	.00046	.00061	-.00776	-.00275	.00071	-.00009	.00649
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9799.0**	**60424.**	**3020.0**
Stddev	11.5	135.	11.0
%RSD	.11760	.22419	.36496
#1	9789.9	60445.	3025.5
#2	9795.1	60547.	3027.2
#3	9812.0	60279.	3007.3

Approved: June 27, 2016

Sample Name: L1606078710 Acquired: 6/25/2016 22:27:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00017	.22720	.00214	.03806	.08552	.00012	52.208	.00022
Stddev	.00078	.00436	.00063	.00014	.00079	.00005	.232	.00016
%RSD	450.39	1.9170	29.562	.37345	.92041	43.173	.44436	71.173
#1	.00025	.22486	.00206	.03796	.08534	.00008	52.380	.00034
#2	.00030	.22452	.00280	.03822	.08638	.00018	52.301	.00028
#3	-.00108	.23223	.00155	.03799	.08484	.00010	51.944	.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00002	.00186	.00266	.23670	.85794	.00848	16.196	.00759
Stddev	.00014	.00018	.00107	.03451	.05555	.00209	.053	.00200
%RSD	849.17	9.9559	40.055	14.580	6.4743	24.705	.32764	26.366
#1	-.00011	.00187	.00273	.22408	.80379	.01054	16.251	.00990
#2	.00017	.00167	.00156	.27574	.91479	.00636	16.192	.00643
#3	-.00001	.00204	.00369	.21027	.85526	.00853	16.145	.00644
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00109	3.4758	.00069	.17816	.00790	-.00197	-.00080	8.3609
Stddev	.00043	.0348	.00145	.00114	.00265	.00100	.00335	.0283
%RSD	39.553	1.0002	208.84	.63768	33.519	50.807	416.39	.33868
#1	.00115	3.5101	.00211	.17685	.00777	-.00150	.00187	8.3874
#2	.00149	3.4766	.00076	.17877	.00531	-.00312	.00028	8.3643
#3	.00063	3.4406	-.00079	.17886	.01060	-.00129	-.00456	8.3311
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078710 Acquired: 6/25/2016 22:27:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00106	.13692	-.00198	-.00343	.00351	.00457	.00538
Stddev	.00037	.00076	.00399	.00152	.00135	.00005	.00158
%RSD	34.945	.55142	201.79	44.378	38.390	1.0573	29.298
#1	.00117	.13773	-.00276	-.00427	.00288	.00452	.00538
#2	.00136	.13624	.00235	-.00167	.00505	.00462	.00381
#3	.00065	.13680	-.00552	-.00433	.00259	.00458	.00696
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9837.0	60854.	3075.5
Stddev	21.0	87.	13.2
%RSD	.21328	.14286	.42972
#1	9813.4	60881.	3070.1
#2	9853.7	60757.	3065.8
#3	9843.8	60924.	3090.5

Approved: June 27, 2016

Sample Name: L1606078718 Acquired: 6/25/2016 22:31:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00018	1.2702	.00086	.07193	.15344	.00016	53.535	.00010
Stddev	.00088	.0133	.00129	.00289	.00068	.00006	.315	.00014
%RSD	504.87	1.0437	151.37	4.0242	.44552	39.145	.58814	143.56
#1	.00078	1.2717	-.00017	.07191	.15299	.00023	53.705	-.00002
#2	-.00096	1.2563	.00043	.07484	.15310	.00014	53.172	.00025
#3	-.00034	1.2827	.00231	.06905	.15423	.00011	53.729	.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00093	.00320	.00404	1.3343	1.2661	.00990	20.655	.10718
Stddev	.00038	.00017	.00073	.0192	.0597	.00249	.173	.00269
%RSD	40.455	5.3029	17.991	1.4379	4.7192	25.172	.83716	2.5057
#1	.00050	.00331	.00423	1.3126	1.1986	.01278	20.828	.10559
#2	.00122	.00328	.00465	1.3416	1.2875	.00842	20.482	.11029
#3	.00107	.00300	.00324	1.3489	1.3123	.00851	20.655	.10568
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00197	9.5571	.00271	.15727	.00713	-.00001	-.00213	9.2004
Stddev	.00025	.0745	.00060	.00312	.00099	.00396	.00180	.0136
%RSD	12.754	.77965	22.028	1.9854	13.840	44119.	84.428	.14793
#1	.00222	9.6099	.00232	.16018	.00745	-.00238	-.00366	9.2098
#2	.00198	9.4719	.00242	.15397	.00602	-.00222	-.00015	9.2066
#3	.00172	9.5895	.00340	.15765	.00792	.00457	-.00260	9.1848
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078718 Acquired: 6/25/2016 22:31:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00007	.21832	.00935	-.00138	.00550	.00807	.00587
Stddev	.00077	.00070	.00368	.00189	.00037	.00008	.00139
%RSD	1050.3	.31961	39.319	137.43	6.8010	.96236	23.773
#1	.00067	.21839	.00655	-.00216	.00508	.00811	.00494
#2	-.00003	.21758	.00798	.00078	.00561	.00812	.00519
#3	-.00086	.21897	.01351	-.00275	.00580	.00798	.00747
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9846.3	60728.	3091.9
Stddev	3.6	208.	32.5
%RSD	.03671	.34247	1.0518
#1	9850.4	60702.	3069.6
#2	9844.9	60535.	3129.2
#3	9843.6	60948.	3076.8

Approved: June 27, 2016

Sample Name: L1606078719 Acquired: 6/25/2016 22:35:34 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00022**	**-.00125**	**.00081**	**-.00052**	**.00163**	**.00015**	**.05700**	**.00032**
Stddev	.00275	.01314	.00222	.00254	.00035	.00003	.05177	.00012
%RSD	1222.4	1048.7	273.08	489.15	21.203	21.890	90.827	38.965
#1	-.00200	-.00415	-.00134	.00209	.00158	.00017	.05365	.00040
#2	.00294	.01310	.00068	-.00066	.00200	.00012	.00698	.00018
#3	-.00162	-.01271	.00310	-.00298	.00131	.00018	.11035	.00038
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00006**	**.00052**	**.00450**	**.02391**	**-.05991**	**-.00020**	**.04697**	**-.00036**
Stddev	.00016	.00162	.00035	.01615	.07304	.00309	.12538	.00213
%RSD	276.80	313.97	7.6965	67.537	121.91	1526.0	266.94	591.72
#1	.00015	.00207	.00475	.04225	-.11673	.00336	-.03849	-.00282
#2	-.00012	.00065	.00464	.01187	-.08546	-.00225	-.01151	.00086
#3	.00015	-.00117	.00410	.01759	.02247	-.00171	.19090	.00088
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00030**	**.25593**	**-.00012**	**.00090**	**.00315**	**-.00362**	**.00327**	**.05573**
Stddev	.00023	.02584	.00048	.00460	.00186	.00138	.00112	.00365
%RSD	74.483	10.098	393.31	510.33	59.051	38.233	34.210	6.5471
#1	.00052	.27908	.00017	.00590	.00100	-.00516	.00452	.05848
#2	.00007	.22805	.00014	-.00005	.00421	-.00248	.00234	.05711
#3	.00032	.26066	-.00068	-.00315	.00425	-.00323	.00296	.05159
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078719 Acquired: 6/25/2016 22:35:34 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00098**	**.00207**	**-.00026**	**-.00547**	**.00052**	**.00525**	**.00241**
Stddev	.00027	.00008	.00293	.00052	.00032	.00008	.00097
%RSD	27.115	4.0421	1136.5	9.5358	61.869	1.5786	40.264
#1	.00096	.00217	-.00025	-.00546	.00057	.00519	.00232
#2	.00073	.00203	-.00319	-.00495	.00081	.00521	.00342
#3	.00126	.00203	.00267	-.00599	.00017	.00534	.00149
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9914.0**	**61492.**	**3065.3**
Stddev	9.7	119.	10.3
%RSD	.09735	.19352	.33672
#1	9905.1	61523.	3061.5
#2	9924.3	61360.	3057.3
#3	9912.7	61592.	3076.9

Approved: June 27, 2016

Sample Name: L1606078709 Acquired: 6/25/2016 22:39:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00063	.01725	.00020	.00148	.00412	.00007	2.5193	.00009
Stddev	.00160	.00300	.00146	.00122	.00055	.00007	.0273	.00025
%RSD	256.33	17.406	733.70	82.744	13.249	105.70	1.0828	269.34
#1	-.00115	.01430	.00189	.00205	.00368	.00003	2.5430	-.00015
#2	.00106	.02030	-.00065	.00007	.00396	.00002	2.4895	.00008
#3	.00197	.01716	-.00064	.00231	.00474	.00015	2.5253	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00035	.00087	.00073	.04349	.03288	.00266	.76988	-.00075
Stddev	.00037	.00063	.00114	.01165	.16835	.00415	.07900	.00106
%RSD	107.35	72.316	156.53	26.779	512.03	155.84	10.261	141.15
#1	.00070	.00029	.00142	.03927	-.06801	.00606	.83027	.00043
#2	-.00004	.00153	.00135	.03454	-.06058	.00390	.79891	-.00163
#3	.00039	.00078	-.00059	.05665	.22723	-.00196	.68048	-.00106
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00016	.20336	.00039	.01881	.00195	-.00218	-.00099	.41497
Stddev	.00052	.02965	.00044	.00286	.00086	.00256	.00440	.00248
%RSD	323.08	14.579	115.44	15.203	44.305	117.44	445.87	.59653
#1	-.00058	.19785	-.00012	.01994	.00207	-.00390	.00409	.41320
#2	.00043	.23538	.00054	.02093	.00103	-.00339	-.00379	.41391
#3	-.00033	.17686	.00073	.01556	.00275	.00076	-.00326	.41779
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078709 Acquired: 6/25/2016 22:39:31 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**.00684**	**.01005**	**.00002**	**.00088**	**.00094**	**.00070**
Stddev	.00032	.00012	.00959	.00200	.00079	.00012	.00147
%RSD	375.67	1.7762	95.434	11093.	88.946	12.687	210.40
#1	.00028	.00687	.01212	-.00222	.00129	.00101	.00212
#2	.00025	.00693	-.00041	.00064	.00138	.00101	.00079
#3	-.00028	.00670	.01844	.00163	-.00002	.00080	-.00081
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10438.**	**64428.**	**3210.2**
Stddev	16.	116.	22.5
%RSD	.15594	.18047	.70191
#1	10456.	64522.	3184.5
#2	10433.	64464.	3226.3
#3	10424.	64298.	3219.8

Approved: June 27, 2016

Sample Name: L1606078711S Acquired: 6/25/2016 22:43:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00901	.25091	.00908	.04716	.02857	.00129	2.7206	.00112
Stddev	.00226	.00893	.00272	.00090	.00098	.00003	.0190	.00018
%RSD	25.135	3.5576	29.953	1.9149	3.4212	2.0076	.69811	15.952
#1	.00654	.24156	.01090	.04731	.02948	.00128	2.7394	.00096
#2	.00949	.25183	.00595	.04798	.02754	.00132	2.7210	.00131
#3	.01099	.25934	.01039	.04620	.02870	.00128	2.7014	.00109
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00484	.01138	.01241	.12881	1.1056	.02562	.95895	.01155
Stddev	.00047	.00095	.00054	.02178	.0733	.00325	.08330	.00186
%RSD	9.6843	8.3365	4.3248	16.912	6.6280	12.671	8.6861	16.074
#1	.00430	.01246	.01288	.12748	1.1086	.02193	.86824	.01328
#2	.00510	.01069	.01253	.10772	1.1774	.02802	1.0320	.00959
#3	.00513	.01099	.01183	.15123	1.0309	.02692	.97661	.01179
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02077	1.4247	.01204	.24038	.01720	.02493	.01091	.53474
Stddev	.00067	.0371	.00126	.00484	.00256	.00194	.00301	.00121
%RSD	3.2317	2.6069	10.481	2.0127	14.893	7.7710	27.534	.22565
#1	.02144	1.3934	.01127	.24412	.01465	.02673	.00768	.53472
#2	.02075	1.4149	.01350	.23492	.01719	.02518	.01362	.53595
#3	.02010	1.4657	.01136	.24211	.01977	.02288	.01143	.53354
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078711S Acquired: 6/25/2016 22:43:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02007**	**.03079**	**.02489**	**.01057**	**.02372**	**.02477**	**.00180**
Stddev	.00062	.00022	.00917	.00037	.00088	.00015	.00112
%RSD	3.0692	.70951	36.838	3.4629	3.6971	.59147	62.402
#1	.02076	.03087	.02809	.01029	.02425	.02492	.00062
#2	.01959	.03097	.01455	.01045	.02271	.02462	.00193
#3	.01984	.03055	.03204	.01098	.02420	.02478	.00285
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10426.**	**64070.**	**3185.9**
Stddev	11.	72.	10.0
%RSD	.10672	.11272	.31382
#1	10417.	64076.	3184.7
#2	10423.	64138.	3176.5
#3	10439.	63994.	3196.4

Approved: June 27, 2016

Sample Name: L1606078712SD Acquired: 6/25/2016 22:47:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-06

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00985	.31033	.01141	.04692	.02793	.00122	2.7514	.00123
Stddev	.00217	.01120	.00035	.00183	.00030	.00009	.0322	.00017
%RSD	21.999	3.6078	3.0998	3.8902	1.0783	7.5455	1.1711	13.538
#1	.00744	.30133	.01139	.04489	.02771	.00131	2.7147	.00104
#2	.01165	.32286	.01106	.04746	.02780	.00124	2.7644	.00130
#3	.01046	.30678	.01177	.04842	.02827	.00112	2.7751	.00135
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00442	.01187	.01326	.16770	1.1233	.02635	1.0271	.01377
Stddev	.00036	.00059	.00082	.01226	.1089	.00034	.0701	.00235
%RSD	8.1823	4.9929	6.1530	7.3088	9.6965	1.2889	6.8269	17.054
#1	.00413	.01121	.01277	.16233	1.0153	.02615	1.0429	.01385
#2	.00482	.01235	.01421	.15905	1.2332	.02617	1.0881	.01607
#3	.00430	.01207	.01281	.18173	1.1215	.02675	.95049	.01138
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.02146	1.3700	.01235	.24640	.01637	.02615	.01276	.67342
Stddev	.00028	.0367	.00082	.00464	.00030	.00374	.00186	.00470
%RSD	1.2904	2.6782	6.6457	1.8816	1.8039	14.304	14.555	.69838
#1	.02172	1.3476	.01220	.24696	.01603	.02416	.01223	.67004
#2	.02148	1.3500	.01162	.25073	.01656	.02383	.01483	.67143
#3	.02117	1.4123	.01324	.24151	.01651	.03046	.01123	.67879
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606078712SD Acquired: 6/25/2016 22:47:19 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment: WG573102-06

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.02141**	**.03084**	**.02074**	**.01089**	**.02357**	**.02510**	**.00112**
Stddev	.00030	.00031	.00329	.00303	.00061	.00018	.00144
%RSD	1.3990	1.0105	15.854	27.843	2.5684	.72342	128.59
#1	.02165	.03055	.02352	.00739	.02416	.02497	-.00034
#2	.02107	.03117	.01711	.01260	.02359	.02501	.00117
#3	.02151	.03078	.02157	.01268	.02295	.02530	.00254
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10421.**	**63887.**	**3174.5**
Stddev	13.	153.	11.8
%RSD	.12017	.23969	.37110
#1	10421.	64063.	3186.8
#2	10434.	63786.	3163.4
#3	10409.	63811.	3173.3

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 22:51:13 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39829	10.014	.39328	.48763	1.0098	.04913	9.9981
Stddev	.00170	.014	.00352	.00156	.0114	.00024	.0222
%RSD	.42599	.13477	.89524	.31896	1.1283	.49183	.22224
#1	.39634	10.025	.39615	.48872	1.0228	.04926	10.024
#2	.39905	9.9992	.38935	.48585	1.0049	.04885	9.9823
#3	.39946	10.019	.39433	.48833	1.0016	.04927	9.9886
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05040	.19854	.49076	.50671	3.9457	49.325	1.0128
Stddev	.00023	.00052	.00245	.00059	.0608	.368	.0109
%RSD	.45580	.25959	.49941	.11597	1.5404	.74655	1.0755
#1	.05026	.19800	.48916	.50736	4.0058	49.743	1.0249
#2	.05027	.19858	.48954	.50622	3.9470	49.185	1.0101
#3	.05066	.19903	.49358	.50654	3.8842	49.047	1.0036
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.4894	.49712	1.0151	50.056	.50217	9.8694	.50095
Stddev	.2168	.00232	.0035	.479	.00145	.0113	.00387
%RSD	2.2847	.46695	.34803	.95643	.28811	.11418	.77216
#1	9.7395	.49968	1.0186	50.588	.50383	9.8565	.50471
#2	9.3739	.49655	1.0152	49.919	.50140	9.8745	.49698
#3	9.3548	.49515	1.0115	49.660	.50126	9.8772	.50116
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 22:51:13 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.1957**	**.39554**	**5.0959**	**.99417**	**1.0077**	**1.0177**	**.50994**
Stddev	.0055	.00372	.0051	.00169	.0107	.0101	.00264
%RSD	.46066	.94108	.09932	.17029	1.0641	.98980	.51770
#1	1.1919	.39655	5.0908	.99223	1.0196	1.0259	.51299
#2	1.2020	.39866	5.0958	.99500	1.0048	1.0064	.50861
#3	1.1932	.39142	5.1010	.99530	.99871	1.0207	.50824
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**.99906**	**1.0000**	F **2.9202**
Stddev	.00128	.0004	.0305
%RSD	.12818	.03891	1.0450
#1	1.0005	1.0005	2.9552
#2	.99795	.99972	2.9065
#3	.99877	.99993	2.8990
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9918.3**	**60256.**	**3072.6**
Stddev	6.2	156.	34.9
%RSD	.06223	.25822	1.1374
#1	9919.7	60085.	3032.6
#2	9923.7	60388.	3088.2
#3	9911.6	60295.	3097.1

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 22:54:51 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00016**	**.00017**	**.00165**	**.00199**	**.00201**	**.00011**	**-.05290**	**-.00006**
Stddev	.00211	.00653	.00321	.00079	.00009	.00008	.03288	.00016
%RSD	1282.1	3936.0	193.87	39.670	4.4194	70.107	62.164	240.48
#1	-.00117	.00347	.00046	.00135	.00191	.00015	-.05612	-.00004
#2	.00226	-.00735	.00528	.00288	.00202	.00015	-.08405	.00008
#3	-.00158	.00438	-.00078	.00176	.00209	.00002	-.01852	-.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00001**	**.00100**	**.00120**	**.03069**	**-.03948**	**.00003**	**-.00536**	**.00015**
Stddev	.00013	.00044	.00061	.00550	.07927	.00292	.09167	.00308
%RSD	995.60	43.627	50.884	17.919	200.80	8764.3	1709.0	2010.9
#1	-.00014	.00145	.00060	.03538	.05200	-.00061	-.10767	.00091
#2	.00000	.00096	.00116	.03205	-.08238	-.00251	.06932	.00278
#3	.00011	.00058	.00182	.02464	-.08805	.00322	.02226	-.00324
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00351**	**.02455**	**-.00057**	**.00388**	**.00014**	**.00562**	**.00087**	**.03540**
Stddev	.00076	.01905	.00055	.00530	.00438	.00343	.00632	.00172
%RSD	21.717	77.584	95.045	136.67	3025.0	61.032	726.47	4.8534
#1	.00266	.03539	-.00120	-.00182	.00487	.00872	.00338	.03589
#2	.00374	.03571	-.00017	.00867	-.00067	.00193	.00554	.03683
#3	.00413	.00256	-.00036	.00480	-.00377	.00622	-.00632	.03349
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 22:54:51 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00099**	**.00045**	**.00606**	**-.00120**	**.00030**	**-.00008**	**.00363**
Stddev	.00065	.00027	.00156	.00179	.00040	.00006	.00038
%RSD	65.956	60.056	25.746	148.69	132.31	80.432	10.554
#1	.00115	.00019	.00786	-.00316	.00076	-.00006	.00364
#2	.00155	.00073	.00529	-.00079	.00014	-.00015	.00325
#3	.00027	.00044	.00504	.00034	.00001	-.00003	.00401
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9678.8**	**59649.**	**2985.7**
Stddev	6.9	105.	9.5
%RSD	.07084	.17603	.31957
#1	9671.2	59575.	2994.0
#2	9681.0	59769.	2975.3
#3	9684.3	59604.	2987.7

Approved: June 27, 2016

Sample Name: PBW LJ Acquired: 6/25/2016 22:58:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00058**	**-.00592**	**.00193**	**-.00086**	**.00070**	**.00018**	**-.07344**	**-.00023**
Stddev	.00178	.00410	.00330	.00167	.00093	.00007	.04096	.00010
%RSD	307.62	69.294	170.45	194.42	133.02	36.643	55.774	43.143
#1	-.00092	-.00625	.00300	-.00003	-.00002	.00015	-.02635	-.00025
#2	-.00217	-.00166	.00456	.00023	.00174	.00026	-.10082	-.00033
#3	.00135	-.00985	-.00176	-.00279	.00036	.00013	-.09313	-.00013
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00016**	**.00069**	**.00074**	**.03692**	**-.07694**	**.00146**	**.06376**	**.00092**
Stddev	.00002	.00038	.00020	.01230	.06836	.00025	.04368	.00067
%RSD	10.723	54.599	26.935	33.312	88.851	16.877	68.506	72.982
#1	-.00017	.00098	.00051	.02894	-.09438	.00132	.11194	.00086
#2	-.00017	.00026	.00082	.03073	-.13489	.00174	.05264	.00161
#3	-.00014	.00084	.00088	.05108	-.00155	.00131	.02672	.00028
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00019**	**.03492**	**.00034**	**.00729**	**.00159**	**-.00194**	**.00113**	**.00052**
Stddev	.00031	.01196	.00046	.00336	.00147	.00263	.00402	.00268
%RSD	159.07	34.258	134.06	46.125	92.547	135.75	357.58	520.37
#1	.00038	.02883	.00039	.00624	.00208	.00024	-.00125	.00071
#2	.00036	.04870	-.00014	.00458	.00275	-.00120	-.00114	.00309
#3	-.00016	.02722	.00078	.01105	-.00006	-.00487	.00577	-.00226
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: PBW LJ Acquired: 6/25/2016 22:58:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00001**	**-.00008**	**.00508**	**-.00273**	**.00114**	**.00044**	**.00066**
Stddev	.00164	.00051	.01008	.00394	.00109	.00015	.00112
%RSD	14590.	648.58	198.46	144.23	96.326	34.462	168.97
#1	.00174	.00051	.01609	.00176	.00157	.00053	.00180
#2	-.00150	-.00033	-.00371	-.00561	-.00011	.00026	-.00044
#3	-.00028	-.00041	.00287	-.00433	.00195	.00052	.00063
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10135.**	**62589.**	**3095.8**
Stddev	24.	186.	5.7
%RSD	.23247	.29671	.18319
#1	10143.	62746.	3101.4
#2	10152.	62637.	3095.9
#3	10108.	62384.	3090.1

Approved: June 27, 2016

Sample Name: LCSW LJ Acquired: 6/25/2016 23:02:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19432**	**4.8961**	**.19607**	**.95390**	**.50477**	**.02378**	**5.0869**	**.02467**
Stddev	.00176	.0169	.00036	.00515	.00157	.00015	.0469	.00006
%RSD	.90778	.34589	.18422	.54033	.31152	.61673	.92299	.23915
#1	.19242	4.8827	.19640	.95171	.50641	.02365	5.1361	.02462
#2	.19591	4.9152	.19569	.95978	.50463	.02394	5.0425	.02466
#3	.19461	4.8905	.19612	.95020	.50328	.02375	5.0822	.02474
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09889**	**.24063**	**.25385**	**1.9878**	**24.436**	**.51652**	**4.7044**	**.24710**
Stddev	.00023	.00063	.00116	.0095	.067	.00203	.0278	.00275
%RSD	.23212	.26116	.45751	.47691	.27355	.39333	.59179	1.1124
#1	.09867	.24099	.25369	1.9863	24.386	.51588	4.7289	.24429
#2	.09913	.24100	.25509	1.9791	24.512	.51880	4.6741	.24978
#3	.09888	.23991	.25278	1.9979	24.411	.51489	4.7100	.24724
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51030**	**25.148**	**.24989**	**4.7689**	**.24945**	**.59139**	**.18766**	**2.4779**
Stddev	.00067	.056	.00070	.0091	.00347	.00400	.01211	.0014
%RSD	.13092	.22440	.27941	.19073	1.3920	.67662	6.4536	.05441
#1	.50970	25.212	.24981	4.7624	.25341	.58680	.17389	2.4770
#2	.51102	25.108	.25063	4.7793	.24803	.59411	.19668	2.4795
#3	.51019	25.124	.24924	4.7649	.24691	.59326	.19239	2.4773
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LCSW LJ Acquired: 6/25/2016 23:02:46 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49047**	**.51082**	**.50961**	**.24925**	**.49196**	**.48864**	**.01928**
Stddev	.00103	.00205	.00391	.00102	.00174	.00019	.00329
%RSD	.20978	.40119	.76774	.40835	.35306	.03834	17.088
#1	.49156	.51042	.51051	.24964	.49257	.48859	.01592
#2	.49032	.50900	.50533	.25000	.49331	.48848	.01940
#3	.48952	.51304	.51300	.24809	.49000	.48885	.02251
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9771.5**	**60209.**	**3050.2**
Stddev	3.0	103.	25.9
%RSD	.03108	.17103	.84982
#1	9769.0	60280.	3044.5
#2	9774.9	60091.	3027.6
#3	9770.6	60258.	3078.5

Approved: June 27, 2016

Sample Name: L1606076404 Acquired: 6/25/2016 23:06:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00050	.02745	.00034	.00861	.00996	.00009	36.605	.00012
Stddev	.00060	.00628	.00209	.00113	.00041	.00005	.105	.00023
%RSD	119.14	22.865	609.72	13.086	4.1637	53.570	.28645	187.24
#1	.00015	.03394	.00014	.00762	.01044	.00014	36.606	.00007
#2	.00016	.02699	-.00164	.00984	.00971	.00010	36.500	.00038
#3	.00120	.02142	.00253	.00837	.00974	.00004	36.710	-.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00009	.01147	.00175	.46715	4.7234	.00625	13.388	.09951
Stddev	.00015	.00054	.00191	.01948	.0341	.00248	.058	.00203
%RSD	171.91	4.6728	109.26	4.1703	.72246	39.751	.43505	2.0444
#1	.00001	.01158	-.00043	.45694	4.6849	.00912	13.373	.10058
#2	-.00027	.01194	.00254	.48961	4.7498	.00477	13.339	.09716
#3	-.00001	.01088	.00314	.45489	4.7356	.00486	13.453	.10078
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00119	34.643	.00396	.03562	.00752	.00339	.00429	4.0453
Stddev	.00028	.078	.00076	.00870	.00150	.00394	.00341	.0166
%RSD	23.933	.22458	19.114	24.436	19.891	116.11	79.423	.40956
#1	.00087	34.732	.00311	.02910	.00894	.00791	.00692	4.0642
#2	.00131	34.587	.00454	.03225	.00766	.00162	.00551	4.0386
#3	.00140	34.610	.00424	.04550	.00596	.00065	.00044	4.0331
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606076404 Acquired: 6/25/2016 23:06:24 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00088	.19364	.00117	-.00333	.00101	.00286	.00160
Stddev	.00025	.00044	.00456	.00363	.00019	.00020	.00097
%RSD	28.325	.22814	391.45	108.96	18.806	7.1675	60.613
#1	.00083	.19387	-.00007	-.00660	.00107	.00306	.00048
#2	.00116	.19313	.00622	-.00396	.00116	.00265	.00224
#3	.00066	.19392	-.00265	.00057	.00080	.00286	.00207
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9727.8	59976.	3053.1
Stddev	23.9	148.	4.4
%RSD	.24547	.24636	.14325
#1	9700.3	60146.	3049.4
#2	9741.5	59877.	3057.9
#3	9741.8	59905.	3052.0

Approved: June 27, 2016

Sample Name: L1606076405S Acquired: 6/25/2016 23:10:17 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19582**	**4.9527**	**.19879**	**.95673**	**.51438**	**.02432**	**42.805**	**.02505**
Stddev	.00335	.0137	.00128	.00425	.00065	.00003	.112	.00030
%RSD	1.7110	.27715	.64557	.44412	.12542	.14170	.26049	1.1988
#1	.19856	4.9529	.19764	.95186	.51480	.02435	42.707	.02530
#2	.19683	4.9664	.19856	.95967	.51471	.02433	42.926	.02513
#3	.19208	4.9389	.20017	.95866	.51364	.02429	42.781	.02472
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09757**	**.25598**	**.24917**	**2.4354**	**29.536**	**.52338**	**18.805**	**.34995**
Stddev	.00078	.00042	.00127	.0181	.040	.00393	.025	.00506
%RSD	.79936	.16278	.51136	.74393	.13605	.74997	.13071	1.4469
#1	.09813	.25630	.25060	2.4222	29.582	.51885	18.805	.34467
#2	.09790	.25551	.24817	2.4278	29.516	.52547	18.830	.35476
#3	.09668	.25612	.24874	2.4560	29.509	.52582	18.780	.35044
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51541**	**61.224**	**.25128**	**4.8968**	**.24942**	**.59855**	**.19087**	**6.6449**
Stddev	.00129	.228	.00073	.0088	.00234	.00299	.00293	.0075
%RSD	.25084	.37164	.28907	.17909	.93699	.50010	1.5365	.11257
#1	.51663	61.311	.25055	4.9006	.24863	.59583	.18976	6.6461
#2	.51554	61.394	.25129	4.9031	.24758	.60176	.19420	6.6517
#3	.51406	60.965	.25200	4.8868	.25205	.59805	.18866	6.6369
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606076405S Acquired: 6/25/2016 23:10:17 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.49233**	**.70070**	**.50542**	**.24860**	**.50000**	**.48511**	**.00505**
Stddev	.00086	.00402	.00572	.00107	.00163	.00066	.00072
%RSD	.17371	.57387	1.1319	.43103	.32683	.13619	14.263
#1	.49184	.70101	.51171	.24742	.49924	.48572	.00441
#2	.49332	.70456	.50403	.24884	.49889	.48521	.00491
#3	.49184	.69653	.50053	.24953	.50188	.48441	.00583
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9652.6**	**59187.**	**3022.5**
Stddev	17.6	118.	14.2
%RSD	.18215	.19875	.46825
#1	9645.4	59224.	3018.6
#2	9639.7	59281.	3010.7
#3	9672.6	59055.	3038.2

Approved: June 27, 2016

Sample Name: L1606076406SD Acquired: 6/25/2016 23:13:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-05

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.19161**	**4.9186**	**.19913**	**.95429**	**.51211**	**.02396**	**42.977**	**.02467**
Stddev	.00161	.0125	.00343	.00448	.00436	.00009	.276	.00017
%RSD	.83957	.25322	1.7203	.46971	.85098	.36242	.64279	.68812
#1	.19148	4.9258	.19753	.95898	.51664	.02396	43.237	.02475
#2	.19328	4.9042	.19680	.95004	.51174	.02387	43.008	.02479
#3	.19007	4.9258	.20306	.95385	.50794	.02404	42.687	.02448
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09706**	**.25178**	**.24789**	**2.4233**	**29.585**	**.51807**	**18.712**	**.35011**
Stddev	.00057	.00068	.00034	.0356	.112	.00327	.182	.00396
%RSD	.58951	.26961	.13796	1.4688	.37990	.63166	.97051	1.1325
#1	.09772	.25103	.24806	2.4608	29.711	.51534	18.898	.35283
#2	.09684	.25235	.24811	2.4189	29.497	.52170	18.704	.34556
#3	.09664	.25196	.24749	2.3900	29.545	.51719	18.535	.35193
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51262**	**61.425**	**.24874**	**4.8787**	**.24757**	**.59558**	**.18865**	**6.6917**
Stddev	.00159	.310	.00133	.0141	.00184	.00348	.00465	.0131
%RSD	.30991	.50524	.53301	.28958	.74220	.58390	2.4654	.19574
#1	.51304	61.773	.24994	4.8746	.24617	.59720	.19309	6.6816
#2	.51396	61.322	.24897	4.8944	.24965	.59795	.18905	6.7065
#3	.51087	61.178	.24732	4.8670	.24689	.59159	.18381	6.6870
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606076406SD Acquired: 6/25/2016 23:13:56 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573484-05

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48585**	**.69992**	**.50428**	**.08491**	**.49039**	**.49499**	**.00464**
Stddev	.00076	.00443	.00835	.00108	.00225	.00372	.00157
%RSD	.15628	.63319	1.6562	1.2751	.45854	.75207	33.818
#1	.48649	.70500	.51389	.08368	.49168	.49197	.00317
#2	.48606	.69789	.49873	.08571	.48779	.49386	.00446
#3	.48501	.69687	.50023	.08535	.49169	.49915	.00630
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9674.4**	**59675.**	**3057.4**
Stddev	5.6	128.	11.6
%RSD	.05826	.21445	.38044
#1	9670.2	59575.	3048.3
#2	9680.8	59819.	3053.4
#3	9672.1	59630.	3070.5

Approved: June 27, 2016

Sample Name: L1606099102 Acquired: 6/25/2016 23:17:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 10 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00035	.08527	.00559	.04109	.01538	.00019	.57467	.00159
Stddev	.00233	.00869	.00189	.00114	.00133	.00004	.01094	.00013
%RSD	663.84	10.192	33.720	2.7694	8.6207	22.830	1.9043	7.9131
#1	.00210	.09496	.00628	.04240	.01605	.00016	.56261	.00167
#2	-.00061	.08272	.00346	.04040	.01386	.00017	.57746	.00145
#3	-.00255	.07815	.00704	.04046	.01624	.00024	.58395	.00167
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00096	.05412	.00582	.38755	1.2527	.00289	.47003	.00071
Stddev	.00015	.00024	.00026	.01093	.0635	.00116	.03189	.00162
%RSD	15.476	.43524	4.4250	2.8204	5.0656	40.155	6.7852	226.02
#1	.00113	.05437	.00590	.40001	1.1922	.00164	.49884	-.00098
#2	.00091	.05406	.00602	.38303	1.3188	.00308	.43576	.00089
#3	.00085	.05391	.00553	.37960	1.2473	.00394	.47550	.00224
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00360	218.01	.00389	.36041	.01046	-.00372	.00482	1.3728
Stddev	.00017	.12	.00068	.00488	.00162	.00083	.00193	.0038
%RSD	4.7149	.05666	17.447	1.3551	15.457	22.416	40.114	.27490
#1	.00375	218.03	.00311	.35572	.01225	-.00460	.00284	1.3685
#2	.00365	218.13	.00423	.36547	.01003	-.00293	.00492	1.3754
#3	.00341	217.88	.00434	.36004	.00911	-.00364	.00670	1.3745
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606099102 Acquired: 6/25/2016 23:17:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 10 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00632**	**.09765**	**.07700**	**-.00067**	**.01524**	**.07845**	**.01879**
Stddev	.00056	.00044	.00219	.00203	.00068	.00014	.00089
%RSD	8.9139	.45291	2.8381	300.30	4.4598	.18118	4.7496
#1	.00609	.09718	.07562	.00166	.01602	.07833	.01918
#2	.00590	.09770	.07952	-.00172	.01485	.07841	.01777
#3	.00696	.09806	.07586	-.00197	.01484	.07861	.01943
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10241.**	**62264.**	**3185.5**
Stddev	15.	196.	4.7
%RSD	.14516	.31405	.14616
#1	10253.	62395.	3189.9
#2	10245.	62357.	3185.8
#3	10224.	62039.	3180.6

Approved: June 27, 2016

Sample Name: L1606099102PS Acquired: 6/25/2016 23:21:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 10 Custom ID2: Custom ID3:
Comment: WG573952-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.20426**	**5.2585**	**.21569**	**1.0545**	**.53374**	**.02552**	**5.8510**	**.02789**
Stddev	.00128	.0258	.00476	.0057	.01070	.00022	.0799	.00005
%RSD	.62816	.49053	2.2048	.54061	2.0041	.85107	1.3656	.17945
#1	.20288	5.2399	.21984	1.0489	.54587	.02533	5.9343	.02786
#2	.20446	5.2478	.21673	1.0543	.52973	.02547	5.7750	.02786
#3	.20543	5.2880	.21050	1.0603	.52563	.02576	5.8438	.02795
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.10258**	**.30280**	**.26705**	**2.4069**	**27.035**	**.52853**	**5.3059**	**.25725**
Stddev	.00041	.00101	.00181	.0614	.546	.00889	.1709	.00709
%RSD	.40425	.33448	.67841	2.5526	2.0183	1.6821	3.2205	2.7577
#1	.10253	.30184	.26742	2.4708	27.658	.53822	5.4655	.26479
#2	.10301	.30272	.26864	2.4018	26.642	.52076	5.1256	.25071
#3	.10219	.30386	.26508	2.3482	26.805	.52661	5.3266	.25626
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.52291**	**252.79**	**.26034**	**5.5827**	**.26863**	**.62508**	**.20920**	**4.0596**
Stddev	.00363	5.13	.00069	.0240	.00232	.00165	.00657	.0117
%RSD	.69330	2.0298	.26564	.43012	.86233	.26363	3.1385	.28905
#1	.52662	258.69	.26066	5.6009	.26650	.62318	.21182	4.0655
#2	.52272	250.27	.26082	5.5917	.26829	.62599	.21406	4.0673
#3	.51938	249.40	.25955	5.5555	.27109	.62608	.20173	4.0461
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606099102PS Acquired: 6/25/2016 23:21:26 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 10 Custom ID2: Custom ID3:
Comment: WG573952-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.51599**	**.62613**	**.61043**	**.25381**	**.52664**	**.60277**	**.02286**
Stddev	.00198	.01306	.01573	.00527	.00286	.00158	.00057
%RSD	.38295	2.0851	2.5770	2.0756	.54391	.26285	2.4777
#1	.51752	.64119	.62824	.25887	.52456	.60412	.02234
#2	.51668	.61903	.59844	.25420	.52545	.60318	.02277
#3	.51376	.61815	.60462	.24835	.52991	.60103	.02347
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9624.7**	**58159.**	**2980.0**
Stddev	48.4	89.	45.7
%RSD	.50312	.15320	1.5350
#1	9582.5	58170.	2928.1
#2	9614.0	58242.	2997.3
#3	9677.6	58064.	3014.5

Approved: June 27, 2016

Sample Name: L1606099102SDL Acquired: 6/25/2016 23:25:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 50 Custom ID2: Custom ID3:
Comment: WG573952-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00117	.01283	.00236	.01119	.00550	.00018	.08026	.00026
Stddev	.00047	.00301	.00204	.00193	.00071	.00008	.03574	.00006
%RSD	40.405	23.474	86.372	17.221	12.910	46.045	44.534	24.105
#1	.00083	.01247	.00021	.01333	.00520	.00027	.04235	.00026
#2	.00098	.01601	.00261	.01064	.00630	.00011	.11335	.00020
#3	.00172	.01002	.00427	.00960	.00498	.00015	.08510	.00032
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00022	.01258	.00098	.10105	.18990	-.00151	.06238	.00028
Stddev	.00006	.00091	.00048	.00767	.08103	.00626	.02555	.00271
%RSD	25.732	7.2135	49.236	7.5891	42.670	413.92	40.953	970.07
#1	.00026	.01275	.00118	.09700	.21903	.00544	.04558	.00279
#2	.00015	.01339	.00043	.10990	.09833	-.00328	.04979	.00064
#3	.00024	.01160	.00132	.09626	.25235	-.00669	.09179	-.00259
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00205	47.703	.00017	.08239	.00525	-.00365	.00219	.31315
Stddev	.00019	.192	.00051	.00485	.00385	.00400	.00870	.00335
%RSD	9.3003	.40144	294.11	5.8889	73.348	109.43	397.40	1.0688
#1	.00183	47.834	.00076	.07694	.00083	-.00754	.00399	.31041
#2	.00217	47.483	-.00009	.08400	.00786	-.00387	-.00727	.31216
#3	.00215	47.792	-.00015	.08624	.00708	.00045	.00985	.31688
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606099102SDL Acquired: 6/25/2016 23:25:03 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 50 Custom ID2: Custom ID3:
Comment: WG573952-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00128**	**.02176**	**.02063**	**.00132**	**.00408**	**.01800**	**.00415**
Stddev	.00053	.00039	.00638	.00109	.00052	.00037	.00045
%RSD	41.627	1.7773	30.936	82.221	12.786	2.0724	10.786
#1	.00158	.02176	.02171	.00142	.00446	.01762	.00425
#2	.00067	.02138	.02640	.00236	.00348	.01802	.00454
#3	.00160	.02215	.01377	.00019	.00429	.01836	.00366
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9804.5**	**59680.**	**2983.6**
Stddev	22.5	103.	15.3
%RSD	.22901	.17203	.51306
#1	9824.1	59682.	2991.0
#2	9809.6	59782.	2993.9
#3	9780.0	59577.	2966.0

Approved: June 27, 2016

Sample Name: L1606099501 Acquired: 6/25/2016 23:29:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 1000 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00086	.00231	.00190	.02254	.12371	.00015	15.528	-.00008
Stddev	.00014	.00712	.00142	.00137	.00163	.00003	.200	.00022
%RSD	16.093	308.09	74.696	6.0976	1.3212	21.126	1.2862	282.14
#1	.00100	.01032	.00271	.02232	.12196	.00014	15.307	.00002
#2	.00072	-.00013	.00272	.02129	.12520	.00012	15.695	.00008
#3	.00085	-.00326	.00026	.02401	.12396	.00018	15.583	-.00034
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00031	.00027	.00136	.08102	.36083	.08036	1.7207	.01281
Stddev	.00017	.00026	.00081	.01723	.08162	.00488	.0426	.00381
%RSD	55.449	94.239	59.356	21.269	22.620	6.0696	2.4776	29.711
#1	.00018	.00004	.00043	.07010	.45158	.07724	1.7124	.01021
#2	.00026	.00024	.00186	.07209	.29342	.07785	1.6828	.01718
#3	.00051	.00055	.00179	.10089	.33750	.08598	1.7669	.01103
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00006	40.169	.00020	.00739	.00147	-.00261	.00048	.01475
Stddev	.00007	.358	.00037	.00186	.00129	.00197	.00359	.00082
%RSD	116.52	.89035	187.93	25.143	87.719	75.441	741.63	5.5637
#1	.00010	39.803	.00022	.00902	.00206	-.00263	-.00044	.01382
#2	.00008	40.518	.00056	.00537	.00236	-.00063	-.00256	.01539
#3	-.00002	40.186	-.00019	.00779	-.00001	-.00456	.00445	.01503
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606099501 Acquired: 6/25/2016 23:29:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 1000 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00059**	**2.0444**	**.00152**	**-.00059**	**-.00017**	**.00315**	**.00110**
Stddev	.00034	.0207	.00598	.00278	.00059	.00019	.00107
%RSD	58.480	1.0109	393.49	469.74	343.26	6.1207	97.322
#1	.00020	2.0228	.00742	-.00167	.00024	.00305	.00162
#2	.00070	2.0639	.00167	-.00267	-.00085	.00303	-.00013
#3	.00086	2.0465	-.00453	.00256	.00009	.00337	.00180
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9773.4**	**59741.**	**3000.7**
Stddev	10.5	113.	31.9
%RSD	.10719	.18996	1.0628
#1	9785.4	59856.	3037.0
#2	9766.5	59737.	2977.2
#3	9768.1	59629.	2987.9

Approved: June 27, 2016

Sample Name: L1606099502 Acquired: 6/25/2016 23:32:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 1000 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00001	-.00479	-.00049	.02240	.12357	.00014	15.278	-.00007
Stddev	.00142	.00423	.00253	.00505	.00148	.00006	.144	.00014
%RSD	22192.	88.307	518.91	22.533	1.1984	44.965	.94397	206.25
#1	.00014	-.00965	-.00218	.02478	.12272	.00011	15.115	-.00014
#2	.00133	-.00281	-.00170	.02581	.12271	.00021	15.329	.00010
#3	-.00149	-.00191	.00242	.01660	.12528	.00010	15.390	-.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	.00065	.00042	.06837	.37746	.07803	1.6184	.01419
Stddev	.00041	.00112	.00027	.00656	.03147	.00401	.0513	.00310
%RSD	83.614	171.25	62.991	9.6001	8.3371	5.1380	3.1674	21.849
#1	.00050	-.00064	.00063	.07323	.35510	.08210	1.6690	.01383
#2	.00090	.00123	.00012	.06091	.36384	.07791	1.6197	.01745
#3	.00007	.00137	.00052	.07098	.41345	.07408	1.5665	.01128
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00036	39.920	-.00037	.00510	.00199	-.00064	.00335	.01167
Stddev	.00036	.366	.00054	.01025	.00130	.00317	.00244	.00126
%RSD	102.18	.91613	144.37	200.85	65.670	494.36	72.931	10.832
#1	-.00077	39.520	-.00039	.00961	.00071	-.00417	.00384	.01302
#2	-.00018	40.239	-.00090	.01233	.00193	.00029	.00550	.01150
#3	-.00011	40.000	.00017	-.00663	.00331	.00196	.00070	.01051
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: L1606099502 Acquired: 6/25/2016 23:32:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 1000 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00055**	**2.0217**	**-.00337**	**-.00055**	**.00074**	**.00315**	**.00087**
Stddev	.00116	.0191	.00559	.00170	.00084	.00020	.00103
%RSD	210.74	.94683	165.64	307.06	114.28	6.4714	118.16
#1	-.00142	2.0010	.00194	-.00241	.00129	.00330	.00185
#2	.00077	2.0388	-.00920	-.00018	-.00023	.00292	.00097
#3	-.00100	2.0253	-.00287	.00093	.00117	.00324	-.00020
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9730.5**	**59619.**	**2989.3**
Stddev	18.2	229.	29.6
%RSD	.18733	.38360	.99126
#1	9750.6	59882.	3021.9
#2	9715.1	59468.	2964.1
#3	9725.8	59506.	2981.9

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 23:36:47 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40930**	**10.238**	**.40819**	**.50228**	**1.0459**	**.04973**	**10.404**
Stddev	.00043	.021	.00043	.00341	.0044	.00005	.020
%RSD	.10460	.20815	.10461	.67892	.41689	.09834	.19571
#1	.40883	10.251	.40851	.50622	1.0496	.04974	10.381
#2	.40944	10.250	.40771	.50027	1.0411	.04968	10.418
#3	.40965	10.214	.40835	.50036	1.0470	.04978	10.414
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.05210**	**.20379**	**.49816**	**.52303**	**4.0622**	**51.067**	**1.0469**
Stddev	.00019	.00042	.00279	.00142	.0372	.160	.0078
%RSD	.35747	.20724	.56062	.27079	.91500	.31299	.74725
#1	.05231	.20342	.50024	.52229	4.1047	51.143	1.0555
#2	.05194	.20371	.49926	.52467	4.0462	50.883	1.0402
#3	.05207	.20425	.49499	.52215	4.0357	51.175	1.0449
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.7214**	**.51746**	**1.0407**	**52.052**	**.51652**	**10.202**	**.51928**
Stddev	.0660	.00152	.0096	.278	.00085	.015	.00145
%RSD	.67863	.29343	.92144	.53407	.16386	.14443	.27838
#1	9.7320	.51900	1.0479	52.343	.51582	10.200	.51894
#2	9.7814	.51597	1.0444	51.789	.51627	10.218	.52086
#3	9.6508	.51742	1.0298	52.022	.51746	10.189	.51803
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 23:36:47 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2277**	**.40544**	**5.2555**	**1.0220**	**1.0464**	**1.0514**	**.52265**
Stddev	.0051	.00303	.0032	.0024	.0054	.0078	.00372
%RSD	.41461	.74796	.06070	.23053	.51974	.73901	.71099
#1	1.2332	.40616	5.2590	1.0221	1.0521	1.0443	.52536
#2	1.2232	.40805	5.2526	1.0196	1.0413	1.0597	.52418
#3	1.2265	.40211	5.2550	1.0243	1.0458	1.0502	.51841
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0117**	**1.0320**	F **2.9933**
Stddev	.0025	.0009	.0109
%RSD	.24458	.08681	.36389
#1	1.0144	1.0330	3.0053
#2	1.0111	1.0312	2.9841
#3	1.0096	1.0318	2.9905
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9691.0**	**59432.**	**3012.0**
Stddev	5.1	44.	12.8
%RSD	.05245	.07397	.42554
#1	9696.2	59404.	2998.2
#2	9686.1	59410.	3023.6
#3	9690.7	59483.	3014.1

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 23:40:25 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00145**	**-.00660**	**-.00143**	**.00292**	**.00177**	**.00016**	**-.01784**	**.00008**
Stddev	.00180	.00657	.00282	.00130	.00082	.00003	.03649	.00007
%RSD	124.44	99.487	197.72	44.383	46.493	20.281	204.59	79.919
#1	.00061	-.01208	-.00468	.00217	.00185	.00015	-.05680	.00016
#2	-.00276	-.00840	.00028	.00442	.00255	.00019	-.01224	.00005
#3	-.00219	.00068	.00012	.00218	.00091	.00013	.01553	.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00005**	**.00038**	**.00052**	**.02154**	**-.09198**	**-.00110**	**-.01322**	**.00018**
Stddev	.00024	.00088	.00138	.01483	.10535	.00141	.06861	.00246
%RSD	523.34	234.01	265.29	68.833	114.54	127.62	519.13	1371.8
#1	-.00015	.00090	-.00012	.01936	.02327	-.00076	.06137	-.00117
#2	-.00022	-.00064	-.00042	.03733	-.18333	-.00265	-.02738	.00302
#3	.00023	.00086	.00210	.00792	-.11589	.00010	-.07365	-.00132
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00348**	**.04821**	**.00016**	**.00585**	**.00443**	**.00314**	**.00156**	**.03553**
Stddev	.00065	.02773	.00095	.00367	.00371	.00275	.00114	.00188
%RSD	18.560	57.517	582.95	62.725	83.852	87.739	73.200	5.2906
#1	.00289	.07448	.00125	.00268	.00440	.00563	.00189	.03713
#2	.00338	.05093	-.00027	.00499	.00073	.00018	.00029	.03346
#3	.00417	.01922	-.00050	.00987	.00816	.00360	.00249	.03599
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 23:40:25 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00025**	**.00051**	**.00604**	**-.00528**	**.00062**	**.00012**	**.00482**
Stddev	.00070	.00006	.00458	.00163	.00029	.00008	.00094
%RSD	280.85	11.978	75.793	30.927	46.562	68.537	19.510
#1	.00047	.00050	.00150	-.00357	.00036	.00006	.00376
#2	.00081	.00057	.01066	-.00682	.00057	.00008	.00556
#3	-.00053	.00045	.00597	-.00546	.00093	.00021	.00513
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9759.5**	**60265.**	**3013.5**
Stddev	10.8	85.	17.4
%RSD	.11040	.14047	.57599
#1	9754.6	60307.	3032.0
#2	9771.9	60320.	2997.5
#3	9752.1	60167.	3011.1

Approved: June 27, 2016

Sample Name: ICSA Acquired: 6/25/2016 23:44:20 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F -.00562	263.10	.00178	-.00976	.00140	.00014	248.10
Stddev	.00117	.16	.00383	.00309	.00113	.00004	.61
%RSD	20.823	.06041	215.50	31.614	80.449	26.067	.24782
#1	-.00542	263.25	-.00154	-.01064	.00266	.00012	248.67
#2	-.00456	262.94	.00597	-.00633	.00047	.00013	247.45
#3	-.00688	263.13	.00090	-.01231	.00109	.00019	248.17
Check ?	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit	.00400						
Low Limit	-.00400						

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00017	-.00060	-.00056	F .01039	95.630	-.04599	.01242
Stddev	.00016	.00022	.00039	.00101	.248	.03226	.00307
%RSD	92.869	36.608	70.420	9.7281	.25928	70.136	24.681
#1	-.00031	-.00080	-.00098	.00928	95.904	-.06101	.01295
#2	-.00020	-.00063	-.00021	.01127	95.420	-.06800	.01519
#3	.00000	-.00037	-.00048	.01062	95.566	-.00896	.00913
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				.00400			
Low Limit				-.00400			

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	236.10	.00050	-.00055	.08182	-.00122	.03425	.00395
Stddev	.41	.00175	.00028	.01035	.00129	.00383	.00558
%RSD	.17524	347.68	51.671	12.652	105.94	11.172	141.13
#1	236.13	.00226	-.00079	.09323	-.00097	.03151	-.00194
#2	235.67	.00051	-.00062	.07920	-.00262	.03262	.00465
#3	236.50	-.00125	-.00024	.07303	-.00007	.03862	.00916
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 27, 2016

Sample Name: ICSA Acquired: 6/25/2016 23:44:20 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00072	-.00782	.00827	.00013	.00094	-.00929	-.00404
Stddev	.00319	.00204	.00198	.00005	.00018	.00784	.00342
%RSD	442.40	26.134	23.931	35.338	18.766	84.463	84.650
#1	.00212	-.01015	.01018	.00009	.00088	-.00188	-.00347
#2	-.00417	-.00696	.00842	.00012	.00080	-.01750	-.00771
#3	-.00012	-.00634	.00622	.00018	.00114	-.00848	-.00094
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	-.00163	-.00560	-.01351
Stddev	.00058	.00018	.00078
%RSD	35.475	3.3032	5.7782
#1	-.00221	-.00565	-.01421
#2	-.00162	-.00574	-.01366
#3	-.00105	-.00539	-.01267
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9256.6	56705.	2976.7
Stddev	13.6	91.	12.8
%RSD	.14727	.16098	.43054
#1	9241.1	56650.	2963.0
#2	9266.9	56811.	2988.4
#3	9261.8	56655.	2978.7

Approved: June 27, 2016

Sample Name: ICSAB Acquired: 6/25/2016 23:48:09 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.52011**	**271.09**	**.26014**	**-.01193**	**.26131**	**.24391**	**254.06**	**.49893**
Stddev	.00309	.31	.00108	.00072	.00207	.00022	1.57	.00035
%RSD	.59470	.11617	.41599	6.0451	.79127	.08930	.61634	.06930
#1	.51910	271.29	.26138	-.01172	.26259	.24374	255.54	.49856
#2	.51765	270.73	.25938	-.01274	.26241	.24383	254.23	.49898
#3	.52358	271.26	.25965	-.01134	.25892	.24415	252.42	.49925
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.24094**	**.25041**	**.26416**	**96.077**	**5.2095**	**.01124**	**238.55**	**.24627**
Stddev	.00056	.00142	.00058	.420	.1352	.00179	1.07	.00329
%RSD	.23299	.56520	.21819	.43687	2.5957	15.928	.44935	1.3378
#1	.24046	.25065	.26387	96.381	5.3478	.01132	239.59	.24346
#2	.24081	.24889	.26378	96.252	5.2032	.00941	238.63	.24545
#3	.24156	.25169	.26482	95.598	5.0776	.01299	237.45	.24990
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00041**	**5.4959**	**.48742**	**.09841**	**.49766**	**.51195**	**.23776**	**.00256**
Stddev	.00052	.0364	.00161	.00635	.00144	.00162	.00501	.00193
%RSD	128.92	.66237	.33105	6.4479	.28893	.31661	2.1076	75.415
#1	-.00068	5.5313	.48557	.09364	.49656	.51344	.24335	.00339
#2	-.00073	5.4979	.48847	.09597	.49929	.51023	.23366	.00393
#3	.00020	5.4586	.48823	.10561	.49712	.51219	.23626	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: ICSAB Acquired: 6/25/2016 23:48:09 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00140**	**.00103**	**-.00323**	**.47273**	**.25455**	**.48271**	**-.01305**
Stddev	.00059	.00027	.00250	.00229	.00097	.00045	.00046
%RSD	41.750	26.302	77.431	.48347	.38190	.09250	3.5206
#1	.00144	.00120	-.00356	.47420	.25352	.48220	-.01343
#2	.00197	.00072	-.00058	.47010	.25467	.48293	-.01317
#3	.00080	.00116	-.00555	.47390	.25546	.48302	-.01254
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9240.5**	**56174.**	**2955.7**
Stddev	9.6	118.	17.7
%RSD	.10383	.20921	.59797
#1	9251.6	56039.	2946.0
#2	9234.1	56253.	2944.9
#3	9235.9	56229.	2976.1

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 23:51:52 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.41108	10.316	.40843	.50327	1.0436	.05024	10.459
Stddev	.00161	.013	.00298	.00077	.0011	.00011	.030
%RSD	.39215	.12124	.73007	.15382	.10203	.22382	.28694
#1	.41119	10.324	.40804	.50264	1.0446	.05015	10.464
#2	.41263	10.302	.40566	.50304	1.0438	.05037	10.487
#3	.40941	10.323	.41158	.50413	1.0425	.05022	10.428
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.05234	.20485	.50201	.52446	4.0759	51.088	1.0484
Stddev	.00020	.00025	.00165	.00123	.0669	.060	.0059
%RSD	.38699	.12298	.32889	.23505	1.6409	.11683	.55819
#1	.05251	.20498	.50023	.52357	4.1126	51.031	1.0491
#2	.05240	.20456	.50349	.52586	3.9987	51.083	1.0423
#3	.05212	.20501	.50230	.52393	4.1164	51.150	1.0539
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	9.8126	.50786	1.0468	51.847	.51713	10.210	.51778
Stddev	.0274	.00264	.0020	.030	.00045	.013	.00098
%RSD	.27960	.52030	.19073	.05848	.08757	.13121	.18906
#1	9.8005	.51090	1.0485	51.876	.51722	10.201	.51858
#2	9.7933	.50648	1.0474	51.816	.51663	10.204	.51807
#3	9.8440	.50619	1.0446	51.848	.51752	10.226	.51669
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Approved: June 27, 2016

Sample Name: CCV Acquired: 6/25/2016 23:51:52 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.2406**	**.41201**	**5.2627**	**1.0255**	**1.0428**	**1.0401**	**.52263**
Stddev	.0022	.00564	.0065	.0018	.0016	.0094	.00276
%RSD	.17501	1.3693	.12412	.17227	.15774	.90409	.52737
#1	1.2431	.41191	5.2553	1.0271	1.0421	1.0459	.52461
#2	1.2392	.40642	5.2651	1.0236	1.0417	1.0292	.52380
#3	1.2396	.41770	5.2677	1.0258	1.0447	1.0451	.51948
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**1.0216**	**1.0330**	F **3.0271**
Stddev	.0027	.0013	.0163
%RSD	.26838	.12831	.53748
#1	1.0227	1.0315	3.0446
#2	1.0236	1.0335	3.0125
#3	1.0185	1.0340	3.0242
Check ?	Chk Pass	Chk Pass	Chk Fail
Value			1.0000
Range			10.000%

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9680.4**	**59250.**	**3000.1**
Stddev	28.5	27.	16.9
%RSD	.29390	.04497	.56388
#1	9711.9	59280.	2982.6
#2	9672.8	59242.	3016.3
#3	9656.6	59229.	3001.5

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 23:55:30 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00125	.00182	.00085	.00036	.00255	.00017	-.02838	-.00012
Stddev	.00082	.00761	.00304	.00225	.00150	.00005	.02296	.00019
%RSD	65.454	418.99	358.52	623.67	58.837	28.075	80.890	157.58
#1	.00189	-.00688	.00358	-.00006	.00362	.00012	-.00325	-.00021
#2	.00153	.00513	.00139	-.00165	.00321	.00020	-.04827	-.00024
#3	.00033	.00720	-.00242	.00279	.00083	.00021	-.03363	.00010
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	.00075	.00113	.02762	-.02056	.00086	-.01380	-.00101
Stddev	.00025	.00070	.00060	.00527	.01321	.00185	.01535	.00114
%RSD	279.31	93.710	52.614	19.071	64.270	216.68	111.21	112.81
#1	-.00016	.00013	.00054	.02156	-.02746	-.00001	-.01846	.00029
#2	.00010	.00151	.00112	.03019	-.02889	-.00040	-.02627	-.00181
#3	.00033	.00061	.00173	.03110	-.00532	.00298	.00334	-.00151
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00326	.05197	-.00006	.00537	-.00040	.00432	.00298	.03440
Stddev	.00030	.03661	.00066	.00611	.00131	.00120	.00482	.00043
%RSD	9.1019	70.433	1169.2	113.80	328.70	27.704	161.98	1.2407
#1	.00299	.03235	.00066	.00591	.00062	.00425	-.00005	.03422
#2	.00322	.09421	-.00018	.01119	-.00187	.00556	.00044	.03489
#3	.00358	.02936	-.00065	-.00099	.00005	.00317	.00854	.03409
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: CCB Acquired: 6/25/2016 23:55:30 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00008**	**.00044**	**.00562**	**-.00278**	**.00100**	**.00010**	**.00427**
Stddev	.00044	.00041	.00677	.00064	.00081	.00020	.00057
%RSD	558.27	93.814	120.49	22.928	80.799	205.25	13.474
#1	.00052	-.00002	.00467	-.00264	.00112	.00024	.00360
#2	.00007	.00078	.01282	-.00348	.00014	-.00013	.00461
#3	-.00036	.00055	-.00063	-.00223	.00174	.00017	.00459
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9746.4**	**60216.**	**3006.7**
Stddev	3.8	76.	7.9
%RSD	.03876	.12554	.26429
#1	9742.9	60301.	3002.9
#2	9746.0	60157.	3001.3
#3	9750.4	60189.	3015.8

Approved: June 27, 2016

Sample Name: LLCCV Acquired: 6/25/2016 23:59:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00916	.16469	.00814	.08024	.01010	.00165	.38196	.00067
Stddev	.00085	.00884	.00209	.00058	.00103	.00003	.03578	.00008
%RSD	9.3027	5.3671	25.632	.71975	10.222	1.6058	9.3673	12.378
#1	.00982	.16059	.01037	.08090	.01024	.00166	.39925	.00061
#2	.00944	.15864	.00624	.07981	.01105	.00162	.40581	.00076
#3	.00820	.17483	.00780	.08002	.00900	.00168	.34082	.00063
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00411	.00454	.00520	.10552	.68660	.08825	.42307	.00800
Stddev	.00033	.00070	.00174	.01827	.09614	.00099	.09575	.00252
%RSD	7.9024	15.364	33.422	17.313	14.002	1.1218	22.633	31.490
#1	.00383	.00377	.00720	.10444	.59286	.08799	.31358	.00534
#2	.00447	.00514	.00437	.12431	.68197	.08934	.49110	.01035
#3	.00404	.00471	.00403	.08782	.78497	.08741	.46454	.00831
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00870	.45736	.01598	.79373	.00972	.08303	.01581	.78238
Stddev	.00060	.01151	.00062	.00779	.00188	.00162	.00507	.00288
%RSD	6.8723	2.5158	3.9084	.98133	19.382	1.9461	32.090	.36823
#1	.00805	.45897	.01600	.80272	.00902	.08401	.02162	.78192
#2	.00924	.44514	.01534	.78909	.01186	.08117	.01226	.78546
#3	.00881	.46798	.01659	.78938	.00829	.08392	.01355	.77975
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 27, 2016

Sample Name: LLCCV Acquired: 6/25/2016 23:59:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v975) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40813**	**.04261**	**.02883**	**.15405**	**.00896**	**.01738**	**.13453**
Stddev	.00156	.00009	.00723	.00166	.00036	.00005	.00136
%RSD	.38342	.20072	25.085	1.0771	4.0551	.28767	1.0139
#1	.40866	.04260	.02755	.15254	.00855	.01741	.13450
#2	.40935	.04253	.02233	.15583	.00907	.01742	.13318
#3	.40636	.04270	.03662	.15377	.00925	.01733	.13591
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9793.5**	**60446.**	**3011.5**
Stddev	14.7	195.	8.9
%RSD	.15029	.32191	.29401
#1	9810.3	60377.	3006.1
#2	9783.4	60295.	3006.7
#3	9786.6	60666.	3021.7

Approved: June 27, 2016

2.3.2 Metals ICP-MS Data

2.3.2.1 Summary Data

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	126F-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/28/2016 13:15
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/28/2016 20:48
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	NI.062816.204859
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.0147		0.00200	0.00100	0.000500
Cadmium, Total	7440-43-9	0.000600	U	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00231	J	0.00400	0.00200	0.00100
Cobalt, Total	7440-48-4	0.0656		0.00200	0.00100	0.000500
Copper, Total	7440-50-8	0.00314	J	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00100	U	0.00200	0.00100	0.000500
Nickel, Total	7440-02-0	0.0313		0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.000584	J	0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0168	J	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	126F-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/29/2016 12:28
Collect Date:	06/14/2016 11:10	**Dilution:**	50	**File ID:**	NI.062916.122812
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Barium, Total	7440-39-3	9.33		0.300	0.150	0.0750
Manganese, Total	7439-96-5	1.07		0.200	0.100	0.0500

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	C08-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/28/2016 13:15
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/28/2016 20:52
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	NI.062816.205210
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00829		0.00200	0.00100	0.000500
Cadmium, Total	7440-43-9	0.000385	J	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00247	J	0.00400	0.00200	0.00100
Cobalt, Total	7440-48-4	0.00683		0.00200	0.00100	0.000500
Copper, Total	7440-50-8	0.00295	J	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00100	U	0.00200	0.00100	0.000500
Nickel, Total	7440-02-0	0.0138		0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.00101	J	0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0250	U	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	C08-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/29/2016 12:31
Collect Date:	06/14/2016 13:10	**Dilution:**	50	**File ID:**	NI.062916.123124
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Barium, Total	7440-39-3	5.12		0.300	0.150	0.0750
Manganese, Total	7439-96-5	3.15		0.200	0.100	0.0500

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	C03-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/28/2016 13:15
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/28/2016 20:55
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	NI.062816.205522
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00684		0.00200	0.00100	0.000500
Cadmium, Total	7440-43-9	0.000600	U	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00200	U	0.00400	0.00200	0.00100
Cobalt, Total	7440-48-4	0.000578	J	0.00200	0.00100	0.000500
Copper, Total	7440-50-8	0.00200	U	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00100	U	0.00200	0.00100	0.000500
Nickel, Total	7440-02-0	0.00400	U	0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.00100	U	0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0153	J	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	C03-061416	**Prep Method:**	3015	**Prep Date:**	06/27/2016 08:30
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574570	**Analyst:**	BKT	**Run Date:**	06/29/2016 12:34
Collect Date:	06/14/2016 14:15	**Dilution:**	50	**File ID:**	NI.062916.123435
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Barium, Total	7440-39-3	1.56		0.300	0.150	0.0750
Manganese, Total	7439-96-5	1.85		0.200	0.100	0.0500

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.3.2.2 QC Summary Data

Example 6020 Calculations
Perkin Elmer NexION 300X

1.0 Initial Calibration (ICAL) Parameters
The system performs linear regression from data consisting of a blank and three standards.

2.0 Calculating the concentration (C) of an element in water using data from prep log, run log, and quantitation report (note:the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (ug/L)	0.1
Vf = Final volume	100
Vi = Initial volume	40
D = Dilution factor as a multiplier (10X = 10)	1
Cx = Concentration of element in (ug/L)	0.25

3.0 Calculating the concentration (C) of an element in soil using data from prep log, run log, and quantitation report (note: the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (ug/L)	0.1
Vf = Final volume	200
Vi = Initial volume	0.5
D = Dilution factor as a multiplier (10X = 10)	1
Cx = Concentration of element in (ug/kg)	40

4.0 Adjusting the concentration to dry weight:

$$Cdry = \frac{Cx \times 100}{Px}$$

Where:	**Example:**
Cx = Concentration calculated as received (wet basis)	40
Px = Percent solids of sample (%wt)	80
$Cdry$ = Concentration calculated as dry weight (ug/kg)	50

50 ug/kg = 0.050 mg/kg

Perkin Elmer NexION ICP/MS

STANDARDS KEY

QC Std 1 - ICV
QC Std 2 - ICB
QC Std 3 - LLICV
QC Std 4 - ICSA
QC Std 5 - ICSAB
QC Std 6 - CCV
QC Std 7 - CCB
QC Std 8 - LLCCV

Calibration Solutions

Analyte	Stock Conc. (mg/L)	S1 (mg/L)	S2 (mg/L)	S3 (mg/L)	S4 (mg/L)
Al	10	0	0.00005	0.05	0.1
Sb	10	0	0.00005	0.05	0.1
As	10	0	0.00005	0.05	0.1
Ba	10	0	0.00005	0.05	0.1
Be	10	0	0.00005	0.05	0.1
Ca	1000	0	0.005	5	10
Cd	10	0	0.0005	0.05	0.1
Cr	10	0	0.0005	0.05	0.1
Co	10	0	0.0005	0.05	0.1
Cu	10	0	0.0005	0.05	0.1
Fe	1000	0	0.005	5	10
Pb	10	0	0.00005	0.05	0.1
Mg	1000	0	0.005	5	10
Mn	10	0	0.00005	0.05	0.1
Ni	10	0	0.00005	0.05	0.1
K	1000	0	0.005	5	10
Se	10	0	0.00005	0.05	0.1
Ag	10	0	0.00005	0.05	0.1
Na	1000	0	0.005	5	10
Tl	10	0	0.00005	0.05	0.1
V	10	0	0.00005	0.05	0.1
U	1000	0	0.00005	0.05	0.1
Zn	10	0	0.00005	0.05	0.1

Microbac Laboratories Inc.
Microwave Digestion Log

Workgroup: WG574242
Analyst: VC
Spike Analyst: VC
Run Date: 06/27/2016 08:30
Method: 3015
Balance: BAL016
Instrument: MW-3
Instrument Start: 06/27/2016 08:39

SOP: ME407 Revison 19
Spike Solution: STD75998
Spike Witness: ERP
HNO3 Lot #: COA18905
40 & 50 ML. DIGESTION TU COA18987
MS Filters- fisher-Lot#R RGT35621

	SAMPLE #	Type	Matrix	Initial Amount	Final Volume	Initial Vessel Wt	Final Vessel Wt	Spike Amount	Due Date
1	WG574242-02	BLANK	1	20 mL	50 mL	182.814 g	182.793 g		
2	WG574242-03	LCS	1	20 mL	50 mL	183.56 g	183.549 g	.25 mL	
3	L16060794-01	SAMP	1	20 mL	50 mL	184.361 g	184.346 g		06/24/16
4	L16060794-02	SAMP	1	20 mL	50 mL	183.398 g	183.379 g		06/24/16
5	L16060794-03	SAMP	1	20 mL	50 mL	184.044 g	184.009 g		06/24/16
6	L16060794-04	SAMP	1	20 mL	50 mL	182.433 g	182.424 g		06/24/16
7	L16060822-02	SAMP	1	20 mL	50 mL	182.575 g	182.563 g		06/24/16
8	L16060822-03	SAMP	1	20 mL	50 mL	182.733 g	182.696 g		06/24/16
9	L16060823-02	SAMP	1	20 mL	50 mL	184.331 g	184.334 g		06/24/16
10	L16060823-03	SAMP	1	20 mL	50 mL	181.954 g	181.917 g		06/24/16
11	L16060823-04	SAMP	1	20 mL	50 mL	184.816 g	184.809 g		06/24/16
12	WG574242-01	REF	2	20 mL	50 mL	183.706 g	183.7 g		
13	L16060839-01	SAMP	2	20 mL	50 mL	183.706 g	183.7 g		06/22/16
14	L16060843-01	SAMP	2	20 mL	50 mL	179.215 g	179.196 g		06/22/16
15	L16060843-02	SAMP	2	20 mL	50 mL	181.896 g	181.872 g		06/22/16
16	L16061410-01	SAMP	1	20 mL	50 mL	181.139 g	181.127 g		07/01/16
17	L16061412-01	SAMP	1	20 mL	50 mL	181.093 g	181.089 g		07/01/16
18	L16061412-02	SAMP	1	20 mL	50 mL	184.474 g	184.466 g		07/01/16
19	L16061413-01	SAMP	1	20 mL	50 mL	181.294 g	181.262 g		07/01/16
20	L16061416-01	SAMP	1	20 mL	50 mL	183.477 g	183.459 g		07/01/16
21	L16061416-02	SAMP	1	20 mL	50 mL	179.518 g	178.772 g		07/01/16
22	L16061421-01	SAMP	1	20 mL	50 mL	183.973 g	183.93 g		07/01/16
23	WG574242-04	DUP	1	20 mL	50 mL	183.873 g	183.86 g		
24	WG574242-05	MS	1	20 mL	50 mL	181.12 g	181.111 g	.25 mL	
25	WG574242-06	MSD	1	20 mL	50 mL	183.984 g	183.969 g	.25 mL	

Analyst: Vicki Collins

Reviewer: Erin Potter

MW_DIG - Modified 09/30/2009
PDF ID: 4825465
Report generated: 06/27/2016 09:45

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062816A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 572184,573475,574522,574570,574571,574573

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	NI.062816.130257	Blank	Blank		1		06/28/16 13:02
2	NI.062816.130609	WG574537-01	Calibration Point		1		06/28/16 13:06
3	NI.062816.130921	WG574537-02	Calibration Point		1		06/28/16 13:09
4	NI.062816.131232	WG574537-03	Calibration Point		1		06/28/16 13:12
5	NI.062816.131544	WG574537-04	Calibration Point		1		06/28/16 13:15
6	NI.062816.131857	WG574537-05	Initial Calibration Verification		1		06/28/16 13:18
7	NI.062816.132210	WG574537-06	Initial Calib Blank		1		06/28/16 13:22
8	NI.062816.132523	WG574537-07	Low Level Initial Calibration V		1		06/28/16 13:25
9	NI.062816.132834	WG574537-08	Interference Check		1		06/28/16 13:28
10	NI.062816.133145	WG574537-09	Interference Check		1		06/28/16 13:31
11	NI.062816.133459	WG574537-10	CCV		1		06/28/16 13:34
12	NI.062816.133810	WG574537-11	CCB		1		06/28/16 13:38
13	NI.062816.134122	L16060362-03	6060227-03		10		06/28/16 13:41
14	NI.062816.134434	L16060362-04	6060227-04		1		06/28/16 13:44
15	NI.062816.134903	L16060362-03	6060227-03	20/50	10		06/28/16 13:49
16	NI.062816.135218	L16060362-04	6060227-04	20/50	1		06/28/16 13:52
17	NI.062816.135925	L16060362-05	6060227-05	20/50	1		06/28/16 13:59
18	NI.062816.140236	L16060362-06	6060227-06	20/50	1		06/28/16 14:02
19	NI.062816.140602	L16060362-07	6060227-07	20/50	1		06/28/16 14:06
20	NI.062816.140914	L16060362-08	6060227-08	20/50	1		06/28/16 14:09
21	NI.062816.141225	L16060362-09	6060227-09	20/50	1		06/28/16 14:12
22	NI.062816.141537	L16060362-10	6060227-10	20/50	1		06/28/16 14:15
23	NI.062816.141850	WG574537-12	CCV		1		06/28/16 14:18
24	NI.062816.142201	WG574537-13	CCB		1		06/28/16 14:22
25	NI.062816.142514	L16060362-11	6060227-11	20/50	1		06/28/16 14:25
26	NI.062816.142825	L16060362-12	6060227-12	20/50	1		06/28/16 14:28
27	NI.062816.143138	WG574537-14	CCV		1		06/28/16 14:31
28	NI.062816.143450	WG574537-15	CCB		1		06/28/16 14:34
29	NI.062816.143803	WG573333-03	Method/Prep Blank	.25/100	1		06/28/16 14:38
30	NI.062816.144114	WG573333-04	Laboratory Control S	.25/100	1		06/28/16 14:41
31	NI.062816.144426	WG573333-02	Reference Sample		1	L16061053-01	06/28/16 14:44
32	NI.062816.144737	WG573333-07	Matrix Spike	.253/100	1	L16061053-01	06/28/16 14:47
33	NI.062816.145049	WG573333-08	Matrix Spike Duplica	.252/100	1	L16061053-01	06/28/16 14:50
34	NI.062816.145400	L16061048-01	TR0062	.252/100	5		06/28/16 14:54

Page: 1 Approved: June 29, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062816A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 572184,573475,574522,574570,574571,574573

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	NI.062816.145711	WG573475-03	Post Digestion Spike		5	L16061048-01	06/28/16 14:57
36	NI.062816.150023	WG573475-04	Serial Dilution		25	L16061048-01	06/28/16 15:00
37	NI.062816.150336	WG574537-16	Interference Check		1		06/28/16 15:03
38	NI.062816.150647	WG574537-17	Interference Check		1		06/28/16 15:06
39	NI.062816.151001	WG574537-18	CCV		1		06/28/16 15:10
40	NI.062816.151312	WG574537-19	CCB		1		06/28/16 15:13
41	NI.062816.151629	WG574537-20	Low Level Continuing Calibra		1		06/28/16 15:16
42	NI.062816.152424	WG573333-01	Reference Sample		10	L16060787-03	06/28/16 15:24
43	NI.062816.152735	WG573333-05	Matrix Spike	.256/100	10	L16060787-03	06/28/16 15:27
44	NI.062816.153046	WG573333-06	Matrix Spike Duplica	.253/100	10	L16060787-03	06/28/16 15:30
45	NI.062816.153358	L16060787-06	U517SS00512A	.256/100	10		06/28/16 15:33
46	NI.062816.153709	L16060787-08	U517SS00612A	.254/100	10		06/28/16 15:37
47	NI.062816.154021	L16060787-13	U517SS00708A	.252/100	10		06/28/16 15:40
48	NI.062816.154332	L16060787-15	U517SS00808A	.257/100	10		06/28/16 15:43
49	NI.062816.154644	L16060787-16	U517SS00808C	.255/100	10		06/28/16 15:46
50	NI.062816.154955	L16060787-17	U517SS00810A	.259/100	10		06/28/16 15:49
51	NI.062816.155308	WG574537-21	CCV		1		06/28/16 15:53
52	NI.062816.155620	WG574537-22	CCB		1		06/28/16 15:56
53	NI.062816.155933	WG573751-04	Method/Prep Blank	20/50	1		06/28/16 15:59
54	NI.062816.160244	WG573751-05	Laboratory Control S	20/50	1		06/28/16 16:02
55	NI.062816.160556	WG573751-02	Reference Sample		1	L16061151-01	06/28/16 16:05
56	NI.062816.160907	WG573751-07	Matrix Spike	20/50	1	L16061151-01	06/28/16 16:09
57	NI.062816.161219	WG573751-08	Matrix Spike Duplica	20/50	1	L16061151-01	06/28/16 16:12
58	NI.062816.161530	WG573751-01	Reference Sample		1	L16060916-02	06/28/16 16:15
59	NI.062816.161841	WG573751-06	Duplicate	20/50	1	L16060916-02	06/28/16 16:18
60	NI.062816.162153	L16060916-01	PHL 1 FRESHWATER IMPO	20/50	1		06/28/16 16:21
61	NI.062816.162504	WG574522-01	Post Digestion Spike		1	L16060916-01	06/28/16 16:25
62	NI.062816.162816	WG574522-02	Serial Dilution		5	L16060916-01	06/28/16 16:28
63	NI.062816.163129	WG574537-23	CCV		1		06/28/16 16:31
64	NI.062816.163441	WG574537-24	CCB		1		06/28/16 16:34
65	NI.062816.163754	WG574522-02	Serial Dilution		25	L16060916-01	06/28/16 16:37
66	NI.062816.164105	L16061142-01	18WW20-062016	20/50	1		06/28/16 16:41
67	NI.062816.164417	L16061142-03	18WW19F-062016	20/50	1		06/28/16 16:44
68	NI.062816.164728	L16061142-05	18WW19FFD-062016	20/50	1		06/28/16 16:47

Page: 2 Approved: June 29, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062816A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 572184,573475,574522,574570,574571,574573

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
69	NI.062816.165039	L16061151-02	LF003SW0311-007 DUP	20/50	1		06/28/16 16:50
70	NI.062816.165351	L16061151-05	LF003SW0312-007	20/50	1		06/28/16 16:53
71	NI.062816.165702	WG573751-03	Reference Sample		1	L16061154-01	06/28/16 16:57
72	NI.062816.170014	WG573751-09	Matrix Spike	20/50	1	L16061154-01	06/28/16 17:00
73	NI.062816.170325	WG573751-10	Matrix Spike Duplica	20/50	1	L16061154-01	06/28/16 17:03
74	NI.062816.170637	L16061154-04	FQC-RB06201601-CB	20/50	1		06/28/16 17:06
75	NI.062816.170950	WG574537-25	CCV		1		06/28/16 17:09
76	NI.062816.171301	WG574537-26	CCB		1		06/28/16 17:13
77	NI.062816.171614	L16061155-01	ST008MW0812-007	20/50	1		06/28/16 17:16
78	NI.062816.171926	L16061155-02	ST008MW0809-007	20/50	1		06/28/16 17:19
79	NI.062816.172237	L16061155-03	ST008MW0816-007	20/50	1		06/28/16 17:22
80	NI.062816.172550	WG574537-27	CCV		1		06/28/16 17:25
81	NI.062816.172902	WG574537-28	CCB		1		06/28/16 17:29
82	NI.062816.184632	L16061142-01	18WW20-062016	20/50	10		06/28/16 18:46
83	NI.062816.184946	WG574537-29	CCV		1		06/28/16 18:49
84	NI.062816.185257	WG574537-30	CCB		1		06/28/16 18:52
85	NI.062816.185610	WG574537-31	Low Level Continuing Calibra		1		06/28/16 18:56
86	NI.062816.195129	WG574242-02	Method/Prep Blank	20/50	1		06/28/16 19:51
87	NI.062816.195440	WG574242-03	Laboratory Control S	20/50	1		06/28/16 19:54
88	NI.062816.195752	WG574242-01	Reference Sample		1	L16060839-01	06/28/16 19:57
89	NI.062816.200103	WG574242-04	Duplicate	20/50	1	L16060839-01	06/28/16 20:01
90	NI.062816.200414	WG574242-05	Matrix Spike	20/50	1	L16060839-01	06/28/16 20:04
91	NI.062816.200726	WG574242-06	Matrix Spike Duplica	20/50	1	L16060839-01	06/28/16 20:07
92	NI.062816.201037	L16060843-01	6060966-01	20/50	1		06/28/16 20:10
93	NI.062816.201349	WG574570-01	Post Digestion Spike		1	L16060843-01	06/28/16 20:13
94	NI.062816.201700	WG574570-02	Serial Dilution		5	L16060843-01	06/28/16 20:17
95	NI.062816.202012	WG574570-02	Serial Dilution		25	L16060843-01	06/28/16 20:20
96	NI.062816.202325	WG574537-32	CCV		1		06/28/16 20:23
97	NI.062816.202637	WG574537-33	CCB		1		06/28/16 20:26
98	NI.062816.202950	L16060794-01	A08-MW06-Y1S2	20/50	1		06/28/16 20:29
99	NI.062816.203301	L16060794-02	A08-MW09-Y1S2	20/50	1		06/28/16 20:33
100	NI.062816.203613	L16060794-03	A09-MW02-Y1S2	20/50	1		06/28/16 20:36
101	NI.062816.203924	L16060794-04	A09-MW07-Y1S2	20/50	1		06/28/16 20:39
102	NI.062816.204236	L16060822-02	18WW03F-061416	20/50	1		06/28/16 20:42

Page: 3 Approved: June 29, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062816A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 572184,573475,574522,574570,574571,574573

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
103	NI.062816.204547	L16060822-03	18CPTMW04SW-061416	20/50	1		06/28/16 20:45
104	NI.062816.204859	L16060823-02	126F-061416	20/50	1		06/28/16 20:48
105	NI.062816.205210	L16060823-03	C08-061416	20/50	1		06/28/16 20:52
106	NI.062816.205522	L16060823-04	C03-061416	20/50	1		06/28/16 20:55
107	NI.062816.205833	L16060843-02	6060966-02	20/50	1		06/28/16 20:58
108	NI.062816.210146	WG574537-34	CCV		1		06/28/16 21:01
109	NI.062816.210457	WG574537-35	CCB		1		06/28/16 21:04
110	NI.062816.210810	L16061410-01	SS007MW0704-007	20/50	1		06/28/16 21:08
111	NI.062816.211122	L16061412-01	SS007MW0B05-007	20/50	1		06/28/16 21:11
112	NI.062816.211433	L16061412-02	SS007MW0B02-007	20/50	1		06/28/16 21:14
113	NI.062816.211745	L16061413-01	DP005MW0504-007	20/50	1		06/28/16 21:17
114	NI.062816.212056	L16061416-01	DP005MW0509-007	20/50	1		06/28/16 21:20
115	NI.062816.212408	L16061416-02	DP005MW0518-007	20/50	1		06/28/16 21:24
116	NI.062816.212719	L16061421-01	LF003MW03	20/50	1		06/28/16 21:27
117	NI.062816.213032	WG574537-36	CCV		1		06/28/16 21:30
118	NI.062816.213343	WG574537-37	CCB		1		06/28/16 21:33
119	NI.062816.213655	WG574205-02	Method/Prep Blank	20/50	1		06/28/16 21:36
120	NI.062816.214007	WG574205-03	Laboratory Control S	20/50	1		06/28/16 21:40
121	NI.062816.214318	WG574205-01	Reference Sample		1	L16061343-01	06/28/16 21:43
122	NI.062816.214630	WG574205-04	Duplicate	20/50	1	L16061343-01	06/28/16 21:46
123	NI.062816.214941	WG574205-05	Matrix Spike	20/50	1	L16061343-01	06/28/16 21:49
124	NI.062816.215253	WG574205-06	Matrix Spike Duplica	20/50	1	L16061343-01	06/28/16 21:52
125	NI.062816.215604	L16060732-01	LH18/24-SP650-6371-GRAB	20/50	1		06/28/16 21:56
126	NI.062816.215915	WG574571-01	Post Digestion Spike		1	L16060732-01	06/28/16 21:59
127	NI.062816.220227	WG574571-02	Serial Dilution		5	L16060732-01	06/28/16 22:02
128	NI.062816.220538	WG574571-02	Serial Dilution		25	L16060732-01	06/28/16 22:05
129	NI.062816.220851	WG574537-38	CCV		1		06/28/16 22:08
130	NI.062816.221202	WG574537-39	CCB		1		06/28/16 22:12
131	NI.062816.221514	L16060732-02	LH18/24-SP650-6371-COMP	20/50	1		06/28/16 22:15
132	NI.062816.221825	L16060762-01	MW-24-001	20/50	1		06/28/16 22:18
133	NI.062816.222137	L16061380-01	XU936TW003A	20/50	1		06/28/16 22:21
134	NI.062816.222448	L16061380-02	XU936TW003C	20/50	1		06/28/16 22:24
135	NI.062816.222800	L16061381-06	EB02	20/50	1		06/28/16 22:28
136	NI.062816.223111	L16061393-01	S6F0627-01	20/50	1		06/28/16 22:31

Page: 4 Approved: June 29, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062816A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 572184,573475,574522,574570,574571,574573

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
137	NI.062816.223423	L16061393-02	S6F0627-02	20/50	1		06/28/16 22:34
138	NI.062816.223734	L16061393-03	S6F0627-03	20/50	1		06/28/16 22:37
139	NI.062816.224046	L16061393-04	S6F0627-04	20/50	1		06/28/16 22:40
140	NI.062816.224357	L16061393-05	S6F0627-05	20/50	1		06/28/16 22:43
141	NI.062816.224710	WG574537-40	CCV		1		06/28/16 22:47
142	NI.062816.225022	WG574537-41	CCB		1		06/28/16 22:50
143	NI.062816.225335	L16061393-06	S6F0627-06	20/50	1		06/28/16 22:53
144	NI.062816.225646	L16061393-07	S6F0627-07	20/50	1		06/28/16 22:56
145	NI.062816.225958	L16061401-01	DP005MW0501-007	20/50	1		06/28/16 22:59
146	NI.062816.230309	L16061402-01	SS007MW0715-007	20/50	1		06/28/16 23:03
147	NI.062816.230621	L16061404-01	DP005MW0503-007	20/50	1		06/28/16 23:06
148	NI.062816.230932	L16061408-01	DP005MW0508-007	20/50	1		06/28/16 23:09
149	NI.062816.231244	L16061409-01	ST008MW0818-007	20/50	1		06/28/16 23:12
150	NI.062816.231556	WG574537-42	CCV		1		06/28/16 23:15
151	NI.062816.231908	WG574537-43	CCB		1		06/28/16 23:19
152	NI.062816.232221	WG574537-44	Low Level Continuing Calibra		1		06/28/16 23:22
153	NI.062816.232534	WG574306-02	Method/Prep Blank	.25/100	1		06/28/16 23:25
154	NI.062816.232845	WG574306-03	Laboratory Control S	.25/100	1		06/28/16 23:28
155	NI.062816.233157	L16061297-07	XU936DS01700A		1	WG574306-01	06/28/16 23:31
156	NI.062816.233508	L16061297-08	XU936DS01700MS	.253/100	1	WG574306-04	06/28/16 23:35
157	NI.062816.233819	L16061297-09	XU936DS01700MSD	.255/100	1	WG574306-05	06/28/16 23:38
158	NI.062816.234131	L16061443-01	GS042QACF035SB062416-E	.256/100	1		06/28/16 23:41
159	NI.062816.234442	L16061443-02	GS042QACF035SB062416-E	.256/100	1		06/28/16 23:44
160	NI.062816.234754	WG574573-01	Post Digestion Spike		1	L16061443-02	06/28/16 23:47
161	NI.062816.235105	WG574573-02	Serial Dilution		5	L16061443-02	06/28/16 23:51
162	NI.062816.235417	WG574573-02	Serial Dilution		25	L16061443-02	06/28/16 23:54
163	NI.062816.235730	WG574537-45	CCV		1		06/28/16 23:57
164	NI.062916.000041	WG574537-46	CCB		1		06/29/16 00:00

Page: 5 Approved: June 29, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	NI.062916.112234	Blank	Blank		1		06/29/16 11:22
2	NI.062916.112545	WG574676-01	Calibration Point		1		06/29/16 11:25
3	NI.062916.112857	WG574676-02	Calibration Point		1		06/29/16 11:28
4	NI.062916.113208	WG574676-03	Calibration Point		1		06/29/16 11:32
5	NI.062916.113520	WG574676-04	Calibration Point		1		06/29/16 11:35
6	NI.062916.113833	WG574676-05	Initial Calibration Verification		1		06/29/16 11:38
7	NI.062916.114146	WG574676-06	Initial Calib Blank		1		06/29/16 11:41
8	NI.062916.114653	WG574676-07	Low Level Initial Calibration V		1		06/29/16 11:46
9	NI.062916.115053	WG574676-08	Interference Check		1		06/29/16 11:50
10	NI.062916.115447	WG574676-09	Interference Check		1		06/29/16 11:54
11	NI.062916.120010	WG574676-10	CCV		1		06/29/16 12:00
12	NI.062916.120734	WG574676-11	CCB		1		06/29/16 12:07
13	NI.062916.121215	L16060661-06	US23SS02502A	.253/100	1		06/29/16 12:12
14	NI.062916.121526	L16060661-07	US23SS02505A	.257/100	1		06/29/16 12:15
15	NI.062916.121838	L16060661-08	US23SS02602A	.257/100	1		06/29/16 12:18
16	NI.062916.122149	L16060661-09	US23SS02605A	.25/100	1		06/29/16 12:21
17	NI.062916.122501	L16060822-03	18CPTMW04SW-061416	20/50	10		06/29/16 12:25
18	NI.062916.122812	L16060823-02	126F-061416	20/50	50		06/29/16 12:28
19	NI.062916.123124	L16060823-03	C08-061416	20/50	50		06/29/16 12:31
20	NI.062916.123435	L16060823-04	C03-061416	20/50	50		06/29/16 12:34
21	NI.062916.123748	WG574676-12	CCV		1		06/29/16 12:37
22	NI.062916.124059	WG574676-13	CCB		1		06/29/16 12:40
23	NI.062916.124412	WG574676-14	Low Level Continuing Calibra		1		06/29/16 12:44
24	NI.062916.125354	WG573548-02	Method/Prep Blank	20/50	1		06/29/16 12:53
25	NI.062916.125705	WG573548-03	Laboratory Control S	20/50	1		06/29/16 12:57
26	NI.062916.130017	WG573548-01	Reference Sample		1	L16060986-23	06/29/16 13:00
27	NI.062916.130328	WG573548-04	Matrix Spike	20/50	1	L16060986-23	06/29/16 13:03
28	NI.062916.130639	WG573548-05	Matrix Spike Duplica	20/50	1	L16060986-23	06/29/16 13:06
29	NI.062916.130951	L16060986-01	MW-07-04	20/50	1		06/29/16 13:09
30	NI.062916.131302	L16060986-03	MW-01-10	20/50	1		06/29/16 13:13
31	NI.062916.131614	WG574572-01	Post Digestion Spike		1	L16060986-03	06/29/16 13:16
32	NI.062916.131925	WG574572-02	Serial Dilution		5	L16060986-03	06/29/16 13:19
33	NI.062916.132237	WG574572-02	Serial Dilution		25	L16060986-03	06/29/16 13:22
34	NI.062916.132550	WG574676-15	CCV		1		06/29/16 13:25

Page: 1 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	NI.062916.132901	WG574676-16	CCB		1		06/29/16 13:29
36	NI.062916.133214	L16060986-05	MW-03-12	20/50	1		06/29/16 13:32
37	NI.062916.133526	L16060986-07	MW-01-07	20/50	1		06/29/16 13:35
38	NI.062916.133837	L16060986-09	MW-03-11	20/50	1		06/29/16 13:38
39	NI.062916.134150	L16060986-11	MW-01-01	20/50	1		06/29/16 13:41
40	NI.062916.134502	L16060986-13	MW-03-09	20/50	1		06/29/16 13:45
41	NI.062916.134813	L16060986-15	MW-01-08	20/50	1		06/29/16 13:48
42	NI.062916.135125	L16060986-17	MW-03-06	20/50	1		06/29/16 13:51
43	NI.062916.135436	L16060986-19	MW-03-05	20/50	1		06/29/16 13:54
44	NI.062916.135748	L16060986-25	MW-01-02	20/50	1		06/29/16 13:57
45	NI.062916.140059	L16060986-27	DUP 1	20/50	1		06/29/16 14:00
46	NI.062916.140412	WG574676-17	CCV		1		06/29/16 14:04
47	NI.062916.140723	WG574676-18	CCB		1		06/29/16 14:07
48	NI.062916.141036	L16060986-29	DUP 2	20/50	1		06/29/16 14:10
49	NI.062916.141347	L16060986-39	EQUIP BLANK 1	20/50	1		06/29/16 14:13
50	NI.062916.141659	L16060986-41	EQUIP BLANK 2	20/50	1		06/29/16 14:16
51	NI.062916.145120	L16061026-02	18WW16F-061616	20/50	1		06/29/16 14:51
52	NI.062916.145431	L16061026-04	129-061616	20/50	1		06/29/16 14:54
53	NI.062916.145742	L16061026-06	120F-061616	20/50	1		06/29/16 14:57
54	NI.062916.150054	L16061026-07	123-061616	20/50	1		06/29/16 15:00
55	NI.062916.150407	WG574676-19	CCV		1		06/29/16 15:04
56	NI.062916.150718	WG574676-20	CCB		1		06/29/16 15:07
57	NI.062916.151032	WG573099-02	Method/Prep Blank	.25/100	1		06/29/16 15:10
58	NI.062916.151343	WG573099-03	Laboratory Control S	.25/100	1		06/29/16 15:13
59	NI.062916.151655	WG573099-01	Reference Sample		10	L16060888-03	06/29/16 15:16
60	NI.062916.152006	WG573099-04	Matrix Spike	.255/100	10	L16060888-03	06/29/16 15:20
61	NI.062916.152318	WG573099-05	Matrix Spike Duplica	.254/100	10	L16060888-03	06/29/16 15:23
62	NI.062916.152629	L16060883-01	WC01	.252/100	1		06/29/16 15:26
63	NI.062916.152941	L16060912-01	WH8011	.25/100	1		06/29/16 15:29
64	NI.062916.153252	WG573717-03	Post Digestion Spike		1	L16060912-01	06/29/16 15:32
65	NI.062916.153604	WG573717-04	Serial Dilution		5	L16060912-01	06/29/16 15:36
66	NI.062916.153915	WG573717-04	Serial Dilution		25	L16060912-01	06/29/16 15:39
67	NI.062916.154228	WG574676-21	CCV		1		06/29/16 15:42
68	NI.062916.154540	WG574676-22	CCB		1		06/29/16 15:45

Page: 2 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
69	NI.062916.154853	WG574001-04	Method/Prep Blank	20/50	1		06/29/16 15:48
70	NI.062916.155204	WG574001-05	Laboratory Control S	20/50	1		06/29/16 15:52
71	NI.062916.155516	WG574001-01	Reference Sample		1	L16061247-02	06/29/16 15:55
72	NI.062916.155827	WG574001-06	Duplicate	20/50	1	L16061247-02	06/29/16 15:58
73	NI.062916.160139	WG574001-02	Reference Sample		1	L16061265-01	06/29/16 16:01
74	NI.062916.160450	WG574001-07	Matrix Spike	20/50	1	L16061265-01	06/29/16 16:04
75	NI.062916.160802	WG574001-08	Matrix Spike Duplica	20/50	1	L16061265-01	06/29/16 16:08
76	NI.062916.161113	L16061247-01	WMB09	20/50	1		06/29/16 16:11
77	NI.062916.161425	WG574688-01	Post Digestion Spike		1	L16061247-01	06/29/16 16:14
78	NI.062916.161736	WG574688-02	Serial Dilution		5	L16061247-01	06/29/16 16:17
79	NI.062916.162049	WG574676-23	CCV		1		06/29/16 16:20
80	NI.062916.162401	WG574676-24	CCB		1		06/29/16 16:24
81	NI.062916.162713	WG574688-02	Serial Dilution		25	L16061247-01	06/29/16 16:27
82	NI.062916.163025	L16061265-04	SS007MWOB16-007	20/50	1		06/29/16 16:30
83	NI.062916.163336	L16061265-05	SS007MWOB21-007	20/50	1		06/29/16 16:33
84	NI.062916.163648	L16061265-06	FQC-RB06211601-CC	20/50	1		06/29/16 16:36
85	NI.062916.163959	L16061266-01	ST008MW0805-007	20/50	1		06/29/16 16:39
86	NI.062916.164310	L16061269-01	FT002MW0204-007	20/50	1		06/29/16 16:43
87	NI.062916.164622	L16061272-01	SS007MW0B18-007	20/50	1		06/29/16 16:46
88	NI.062916.164933	L16061272-02	SS007MWN752-007	20/50	1		06/29/16 16:49
89	NI.062916.165245	L16061279-01	XU936TW007A	20/50	1		06/29/16 16:52
90	NI.062916.165556	L16061279-01	XU936TW007A	20/50	50		06/29/16 16:55
91	NI.062916.165909	WG574676-25	CCV		1		06/29/16 16:59
92	NI.062916.170220	WG574676-26	CCB		1		06/29/16 17:02
93	NI.062916.170533	WG574001-03	Reference Sample		1	L16061279-02	06/29/16 17:05
94	NI.062916.170845	WG574001-03	Reference Sample		50	L16061279-02	06/29/16 17:08
95	NI.062916.171156	WG574001-09	Matrix Spike	20/50	1	L16061279-02	06/29/16 17:11
96	NI.062916.171508	WG574001-09	Matrix Spike	20/50	50	L16061279-02	06/29/16 17:15
97	NI.062916.171819	WG574001-10	Matrix Spike Duplica	20/50	1	L16061279-02	06/29/16 17:18
98	NI.062916.172131	WG574001-10	Matrix Spike Duplica	20/50	50	L16061279-02	06/29/16 17:21
99	NI.062916.172442	L16061279-05	XU936TW019A	20/50	1		06/29/16 17:24
100	NI.062916.172754	L16061279-05	XU936TW019A	20/50	50		06/29/16 17:27
101	NI.062916.173105	L16061281-01	J6F0803-01	20/50	1		06/29/16 17:31
102	NI.062916.173419	WG574676-27	CCV		1		06/29/16 17:34

Page: 3 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
103	NI.062916.173730	WG574676-28	CCB		1		06/29/16 17:37
104	NI.062916.174043	L16061294-03	EB01	20/50	1		06/29/16 17:40
105	NI.062916.174355	L16061294-03	EB01	20/50	50		06/29/16 17:43
106	NI.062916.174706	L16061296-01	OUTFALL 002/COMP	20/50	1		06/29/16 17:47
107	NI.062916.175019	WG574676-29	Interference Check		1		06/29/16 17:50
108	NI.062916.175331	WG574676-30	Interference Check		1		06/29/16 17:53
109	NI.062916.175644	WG574676-31	CCV		1		06/29/16 17:56
110	NI.062916.175955	WG574676-32	CCB		1		06/29/16 17:59
111	NI.062916.180308	WG574676-33	Low Level Continuing Calibra		1		06/29/16 18:03
112	NI.062916.180621	WG574192-02	Method/Prep Blank	.25/100	1		06/29/16 18:06
113	NI.062916.180932	WG574192-03	Laboratory Control S	.25/100	1		06/29/16 18:09
114	NI.062916.181244	WG574192-01	Reference Sample		1	L16061206-01	06/29/16 18:12
115	NI.062916.181555	L16061206-02	LF003SD0311-007 DUP	.251/100	1		06/29/16 18:15
116	NI.062916.181907	WG574192-04	Matrix Spike	.25/100	1	L16061206-01	06/29/16 18:19
117	NI.062916.182218	WG574192-05	Matrix Spike Duplica	.253/100	1	L16061206-01	06/29/16 18:22
118	NI.062916.182530	L16061206-05	LF003SD0312-007	.25/100	1		06/29/16 18:25
119	NI.062916.182841	WG574690-01	Post Digestion Spike		1	L16061206-05	06/29/16 18:28
120	NI.062916.183153	WG574690-02	Serial Dilution		5	L16061206-05	06/29/16 18:31
121	NI.062916.183504	WG574690-02	Serial Dilution		25	L16061206-05	06/29/16 18:35
122	NI.062916.183818	WG574676-34	CCV		1		06/29/16 18:38
123	NI.062916.184129	WG574676-35	CCB		1		06/29/16 18:41
124	NI.062916.184442	L16061207-01	SS007SD0709-007	.251/100	1		06/29/16 18:44
125	NI.062916.184754	L16061207-02	SS007SD0710-007	.25/100	1		06/29/16 18:47
126	NI.062916.185105	L16061208-01	DP005SD0513-007	.252/100	1		06/29/16 18:51
127	NI.062916.185417	L16061208-02	DP005SD0514-007	.251/100	1		06/29/16 18:54
128	NI.062916.185728	L16061291-01	V6F0270-01	.253/100	1		06/29/16 18:57
129	NI.062916.190040	L16061381-01	XU936DS00500A	.254/100	1		06/29/16 19:00
130	NI.062916.190351	L16061381-01	XU936DS00500A	.254/100	50		06/29/16 19:03
131	NI.062916.190703	L16061381-02	XU936DS00505A	.258/100	1		06/29/16 19:07
132	NI.062916.191014	L16061381-02	XU936DS00505A	.258/100	50		06/29/16 19:10
133	NI.062916.191327	WG574676-36	CCV		1		06/29/16 19:13
134	NI.062916.191639	WG574676-37	CCB		1		06/29/16 19:16
135	NI.062916.191952	L16061381-03	XU936DS00300A	.252/100	1		06/29/16 19:19
136	NI.062916.192303	L16061381-03	XU936DS00300A	.252/100	50		06/29/16 19:23

Page: 4 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP

Analyst1: BKT Analyst2: N/A

Method: 6020/6020A/200.8 SOP: ME700A Rev: 2

Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705

ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607

CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345

Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
137	NI.062916.192615	L16061381-04	XU936DS00305A	.25/100	1		06/29/16 19:26
138	NI.062916.192926	L16061381-04	XU936DS00305A	.25/100	50		06/29/16 19:29
139	NI.062916.193237	L16061381-05	XU936DS00305C	.253/100	1		06/29/16 19:32
140	NI.062916.193550	L16061381-05	XU936DS00305C	.253/100	50		06/29/16 19:35
141	NI.062916.193901	L16061381-07	XU936DS00400A	.252/100	1		06/29/16 19:39
142	NI.062916.194213	L16061381-07	XU936DS00400A	.252/100	50		06/29/16 19:42
143	NI.062916.194524	L16061384-02	BFB-16-019	.253/100	1		06/29/16 19:45
144	NI.062916.194835	L16061384-04	BFB-16-020	.253/100	1		06/29/16 19:48
145	NI.062916.195149	WG574676-38	CCV		1		06/29/16 19:51
146	NI.062916.195500	WG574676-39	CCB		1		06/29/16 19:55
147	NI.062916.195813	L16061384-06	BFB-16-021	.253/100	1		06/29/16 19:58
148	NI.062916.200124	L16061384-08	BFB-16-022	.251/100	1		06/29/16 20:01
149	NI.062916.200436	L16061384-10	BFB-16-023	.254/100	1		06/29/16 20:04
150	NI.062916.200747	L16061384-12	BFB-16-024	.254/100	1		06/29/16 20:07
151	NI.062916.201101	WG574676-40	CCV		1		06/29/16 20:11
152	NI.062916.201412	WG574676-41	CCB		1		06/29/16 20:14
153	NI.062916.201726	WG573291-03	Method/Prep Blank	20/50	1		06/29/16 20:17
154	NI.062916.202037	WG573291-04	Laboratory Control S	20/50	1		06/29/16 20:20
155	NI.062916.202348	WG573291-01	Reference Sample		1	L16060974-01	06/29/16 20:23
156	NI.062916.202700	WG573291-05	Matrix Spike	20/50	1	L16060974-01	06/29/16 20:27
157	NI.062916.203011	WG573291-06	Matrix Spike Duplica	20/50	1	L16060974-01	06/29/16 20:30
158	NI.062916.203323	WG573790-03	Post Digestion Spike		1	L16060974-01	06/29/16 20:33
159	NI.062916.203634	WG573790-04	Serial Dilution		5	L16060974-01	06/29/16 20:36
160	NI.062916.203946	WG573790-04	Serial Dilution		25	L16060974-01	06/29/16 20:39
161	NI.062916.204259	WG574676-42	CCV		1		06/29/16 20:42
162	NI.062916.204611	WG574676-43	CCB		1		06/29/16 20:46

Page: 5 Approved: July 05, 2016

Microbac Laboratories Inc.

Data Checklist

Date: 28-JUN-2016
Analyst: BKT
Analyst: NA
Method: 6020/6020A/200.8
Instrument: ICP-MS2
Curve Workgroup: 574537
Runlog ID: 75986
Analytical Workgroups: 572184,573475,574522,574570,574571,574573

STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	
ICB/CCB	X
ICSA/ICSAB	X
CRI	
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	X
Client Forms	X
Level X	
Level 3	762,1380,1381
Level 4	732,787,822,823,1048,1142,1151,1154
Level 4	1155,1401,1402,1404,1408,1409,1410
Level 4	1412,1413,1416,1421,1443
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	BKT
Secondary Reviewer	KHR
Comments	

Primary Reviewer:
29-JUN-2016

Secondary Reviewer:
29-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-29-2016 15:32:08

Microbac Laboratories Inc.

Data Checklist

Date: 29-JUN-2016
Analyst: BKT
Analyst: NA
Method: 6020/6020A/200.8
Instrument: ICP-MS2
Curve Workgroup: 574676
Runlog ID: 76013
Analytical Workgroups: 573722,574570,574572,573717,574688,574690,573790

STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	
ICB/CCB	X
ICSA/ICSAB	X
CRI	
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	X
Client Forms	X
Level X	
Level 3	1279,1294,1381,1384
Level 4	661,822,823,883,912,974,986,1026
Level 4	1206,1207,1208,1265,1266,1269,1272
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	BKT
Secondary Reviewer	JYH
Comments	

Primary Reviewer:
04-JUL-2016

Secondary Reviewer:
05-JUL-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUL-05-2016 12:36:33

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 6020A

AAB#: WG574570

Login Number: L16060823

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
126F-061416	02	06/14/16					06/27/2016	12.9	180		06/29/16	15.1	180	
126F-061416	02	06/14/16					06/27/2016	12.9	180		06/28/16	14.4	180	
C08-061416	03	06/14/16					06/27/2016	12.8	180		06/28/16	14.3	180	
C08-061416	03	06/14/16					06/27/2016	12.8	180		06/29/16	15	180	
C03-061416	04	06/14/16					06/27/2016	12.8	180		06/28/16	14.3	180	
C03-061416	04	06/14/16					06/27/2016	12.8	180		06/29/16	14.9	180	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4829709
Report generated 06/30/2016 09:23

METHOD BLANK SUMMARY

Login Number: L16060823
Blank File ID: NI.062816.195129
Prep Date: 06/27/16 08:30
Analyzed Date: 06/28/16 19:51
Analyst: BKT

Work Group: WG574570
Blank Sample ID: WG574242-02
Instrument ID: ICP-MS2
Method: 6020A

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG574242-03	NI.062816.195440	06/28/16 19:54	01
DUP	WG574242-04	NI.062816.200103	06/28/16 20:01	01
126F-061416	L16060823-02	NI.062816.204859	06/28/16 20:48	02
C08-061416	L16060823-03	NI.062816.205210	06/28/16 20:52	02
C03-061416	L16060823-04	NI.062816.205522	06/28/16 20:55	02
126F-061416	L16060823-02	NI.062916.122812	06/29/16 12:28	DL01
C08-061416	L16060823-03	NI.062916.123124	06/29/16 12:31	DL01
C03-061416	L16060823-04	NI.062916.123435	06/29/16 12:34	DL01

Report Name: BLANK_SUMMARY
PDF File ID: 4829710
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060823 Prep Date: 06/27/16 08:30 Sample ID: WG574242-02
Instrument ID: ICP-MS2 Run Date: 06/28/16 19:51 Prep Method: 3015
File ID: NI.062816.195129 Analyst: BKT Method: 6020A
Workgroup (AAB#): WG574570 Matrix: Water Units: mg/L
Contract #: Cal ID: ICP-MS-28-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Antimony, Total	0.000500	0.00200	0.000500	1	U
Arsenic, Total	0.000500	0.00200	0.000500	1	U
Barium, Total	0.00150	0.00600	0.00150	1	U
Cadmium, Total	0.000300	0.00120	0.000300	1	U
Chromium, Total	0.00100	0.00400	0.00100	1	U
Cobalt, Total	0.000500	0.00200	0.000500	1	U
Copper, Total	0.00100	0.00400	0.00100	1	U
Lead, Total	0.000500	0.00200	0.000500	1	U
Manganese, Total	0.00100	0.00400	0.00100	1	U
Nickel, Total	0.00200	0.00800	0.00200	1	U
Silver, Total	0.000500	0.00200	0.000500	1	U
Thallium, Total	0.000100	0.000400	0.000100	1	U
Vanadium, Total	0.000500	0.00200	0.000500	1	U
Zinc, Total	0.0125	0.0500	0.0125	1	U

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4829711
29-JUN-2016 08:34

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060823 Run Date: 06/28/2016 Sample ID: WG574242-03
Instrument ID: ICP-MS2 Run Time: 19:54 Prep Method: 3015
File ID: NI.062816.195440 Analyst: BKT Method: 6020A
Workgroup (AAB#): WG574570 Matrix: Water Units: mg/L
QC Key: DOD4 Lot#: STD75998 Cal ID: ICP-MS-28-JUN-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Antimony, Total	0.125	0.120	96.4	80 - 120	
Arsenic, Total	0.125	0.124	99.0	80 - 120	
Barium, Total	0.125	0.121	96.5	80 - 120	
Cadmium, Total	0.125	0.120	95.7	80 - 120	
Chromium, Total	0.125	0.125	100	80 - 120	
Cobalt, Total	0.125	0.122	97.6	80 - 120	
Copper, Total	0.125	0.127	102	80 - 120	
Lead, Total	0.125	0.125	99.8	80 - 120	
Manganese, Total	0.125	0.128	103	80 - 120	
Nickel, Total	0.125	0.126	100	80 - 120	
Silver, Total	0.125	0.123	98.2	80 - 120	
Thallium, Total	0.125	0.122	97.3	80 - 120	
Vanadium, Total	0.125	0.123	98.1	80 - 120	
Zinc, Total	0.125	0.129	103	80 - 120	

LCS - Modified 03/06/2008
PDF File ID: 4829712
Report generated: 06/29/2016 08:34

Microbac Laboratories Inc.
MATRIX SPIKE AND MATRIX SPIKE DUP (MS/MSD)

Loginnum: L16060823 Cal ID: ICP-MS2- Worknum: WG574570
Instrument ID: ICP-MS2 Contract #: Method: 6020A
Parent ID: WG574242-01 File ID: NI.062816.195752 Dil: 1 Matrix: WATER
Sample ID: WG574242-05 MS File ID: NI.062816.200414 Dil: 1 Units: mg/L
Sample ID: WG574242-06 MSD File ID: NI.062816.200726 Dil: 1

Analyte	Parent	MS Spiked	MS Found	MS %Rec	MSD Spiked	MSD Found	MSD %Rec	%RPD	%Rec Limits	RPD Limit	Q
Antimony	0.000998	0.125	0.123	97.9	0.125	0.120	95.5	2.49	80 - 120	20	
Arsenic	0.00141	0.125	0.127	101	0.125	0.126	99.9	0.561	80 - 120	20	
Barium	0.0414	0.125	0.160	95.3	0.125	0.156	92.0	2.54	80 - 120	20	
Cadmium	ND	0.125	0.121	97.2	0.125	0.118	94.0	3.31	80 - 120	20	
Chromium	ND	0.125	0.121	96.6	0.125	0.124	99.2	2.65	80 - 120	20	
Cobalt	ND	0.125	0.121	97.2	0.125	0.120	96.3	0.931	80 - 120	20	
Copper	0.00439	0.125	0.124	95.7	0.125	0.125	96.4	0.655	80 - 120	20	
Lead	0.00115	0.125	0.126	100	0.125	0.124	98.6	1.64	80 - 120	20	
Manganese	0.00611	0.125	0.129	98.4	0.125	0.129	98.3	0.106	80 - 120	20	
Nickel	0.00410	0.125	0.123	95.0	0.125	0.125	96.9	1.96	80 - 120	20	
Silver	ND	0.125	0.122	97.6	0.125	0.118	94.3	3.46	80 - 120	20	
Thallium	ND	0.125	0.123	98.5	0.125	0.121	97.0	1.54	80 - 120	20	
Vanadium	0.00107	0.125	0.124	98.5	0.125	0.123	97.8	0.621	80 - 120	20	
Zinc	0.0337	0.125	0.154	96.2	0.125	0.156	97.7	1.20	80 - 120	20	

* FAILS %REC LIMIT

FAILS RPD LIMIT

NOTE: This is an internal quality control sample.

WG_MS_MSD_DRYWT - Modified 05/26/2011
PDF File ID: 4829713
Report generated 06/29/2016 08:34

Microbac Laboratories Inc.
Serial Dilution Report

Login: L16060823 **Worknum:** WG574570
Instrument: ICP-MS2 **Method:** 6020A
Serial Dil: WG574570-02 **File ID:** NI.062816.201700 **Dil:** 5 **Units:** ug/L
Sample: L16060843-01 **File ID:** NI.062816.201037 **Dil:** 1

Analyte	Sample	Qual	Serial Dil	Qual	% Diff	Q
Antimony	0.382	F	2.24	F	486.00	
Arsenic	ND	U	ND	U		
Barium	29.9	X	30.7	X	2.82	
Cadmium	0.151	F	ND	U		
Chromium	ND	U	ND	U		
Cobalt	7.40	X	7.21	X	2.48	
Copper	0.509	F	ND	U		
Lead	ND	U	ND	U		
Manganese	716		689		3.69	
Nickel	17.5	X	17.0	X	2.69	
Silver	ND	U	ND	U		
Thallium	ND	U	ND	U		
Vanadium	ND	U	ND	U		
Zinc	55.5	X	55.3	F	0.51	

U = Result is below MDL.
F = Result is greater than or equal to MDL and less than the RL.
X = Result is greater than or equal to RL and less than 100 times the MDL.
E = %D exceeds control limit of 10% and initial sample result is greater than or equal to 100 times the MDL.

SERIAL_DIL - Modified 09/22/2008
PDF File ID: 4829707
06/29/2016 08:34

Microbac Laboratories Inc.
POST SPIKE REPORT

Sample Login ID: L16060823 **Worknum:** WG574570
Instrument ID: ICP-MS2 **Method:** 6020A
Post Spike ID: WG574570-01 **File ID:** NI.062816.201349 **Dil:** 1 **Units:** ug/L
Sample ID: L16060843-01 **File ID:** NI.062816.201037 **Dil:** 1 **Matrix:** Water

Analyte	Post Spike Result	C	Sample Result	C	Spike Added(SA)	% R	Control Limit %R	Q
ANTIMONY	49.7		0.382	F	50	98.7	75 - 125	
ARSENIC	51.2		0	U	50	102.5	75 - 125	
BARIUM	78.1		29.9		50	96.4	75 - 125	
CADMIUM	49.2		0.151	F	50	98.1	75 - 125	
CHROMIUM	50.1		0	U	50	100.1	75 - 125	
COBALT	59.3		7.40		50	103.7	75 - 125	
COPPER	50.9		0.509	F	50	100.8	75 - 125	
LEAD	52.4		0	U	50	104.7	75 - 125	
MANGANESE	748		716		50	63.6	75 - 125	N
NICKEL	66.6		17.5		50	98.2	75 - 125	
SILVER	48.6		0	U	50	97.3	75 - 125	
THALLIUM	50.9		0	U	50	101.8	75 - 125	
VANADIUM	50.8		0	U	50	101.7	75 - 125	
ZINC	115		55.5		50	119.5	75 - 125	

N = % Recovery exceeds control limits
F = Result is between MDL and RL
U = Sample result is below MDL. A value of zero is used in the calculation

POST_SPIKE - Modified 03/06/2008
PDF File ID: 4829708
Report generated: 06/29/2016 08:34

Microbac Laboratories Inc.
Initial Calibration Summary

Login:	L16060823	Workgroup (AAB#):	WG574570
Analytical Method:	6020A	Instrument ID:	ICP-MS2
ICAL Worknum:	WG574537	Initial Calibration Date:	28-JUN-2016 13:15

	WG574537-01		WG574537-02		WG574537-03		WG574537-04			
	Conc	INT	Conc	INT	Conc	INT	Conc	INT	R	Q
ANTIMONY	0	47.7	.4	303	50	245000	100	482000	.999953	
ARSENIC	0	-33.8	.4	11.1	50	59300	100	117000	.999972	
BARIUM	0	33.7	.4	159	50	107000	100	212000	.999991	
CADMIUM	0	3.30	.4	116	50	110000	100	219000	.999985	
CHROMIUM	0	7410	.4	8070	50	376000	100	727000	.999989	
COBALT	0	227	.4	599	50	370000	100	731000	.999947	
COPPER	0	135	.4	280	50	102000	100	195000	.999984	
LEAD	0	933	.4	2130	50	978000	100	1950000	.999942	
MANGANESE	0	835	.4	1350	50	340000	100	679000	.99986	
NICKEL	0	183	.4	306	50	100000	100	194000	1	
SILVER	0	79.0	.4	455	50	352000	100	679000	.999803	
THALLIUM	0	5.30	.4	473	50	438000	100	866000	.999971	
VANADIUM	0	1490	.4	1870	50	370000	100	719000	.999996	
ZINC	0	520	.4	1010	50	59300	100	114000	.999999	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
*** = Out of Compliance; R < 0.995**

INT_CAL_ICP - Modified 03/06/2008
PDF File ID: 4829716
Report generated: 30-JUN-2016 09:24

Microbac Laboratories Inc.
Initial Calibration Summary

Login:	L16060823	Workgroup (AAB#):	WG574570
Analytical Method:	6020A	Instrument ID:	ICP-MS2
ICAL Worknum:	WG574676	Initial Calibration Date:	29-JUN-2016 11:35

	WG574676-01		WG574676-02		WG574676-03		WG574676-04			
	Conc	INT	Conc	INT	Conc	INT	Conc	INT	R	Q
ANTIMONY	0	28.2	.4	255	50	227000	100	448000	.999981	
ARSENIC	0	-40.0	.4	-10.4	50	57000	100	110000	.999904	
BARIUM	0	25.3	.4	136	50	99600	100	196000	.999974	
CADMIUM	0	1.30	.4	122	50	102000	100	200000	.999946	
CHROMIUM	0	7540	.4	7930	50	379000	100	748000	.999997	
COBALT	0	239	.4	575	50	337000	100	648000	.99988	
COPPER	0	103	.4	285	50	94300	100	184000	.999979	
LEAD	0	873	.4	2280	50	926000	100	1830000	.999949	
MANGANESE	0	784	.4	1380	50	355000	100	686000	.999925	
NICKEL	0	221	.4	324	50	94600	100	186000	.999991	
SILVER	0	86.3	.4	412	50	318000	100	624000	.999961	
THALLIUM	0	1.70	.4	417	50	406000	100	804000	.999963	
VANADIUM	0	1430	.4	1850	50	385000	100	745000	.999935	
ZINC	0	365	.4	549	50	56000	100	108000	.999932	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
*** = Out of Compliance; R < 0.995**

INT_CAL_ICP - Modified 03/06/2008
PDF File ID: 4829716
Report generated: 30-JUN-2016 09:24

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16060823
Instrument ID: ICP-MS2
File ID: NI.062816.132210
Workgroup (AAB#): WG574570
Matrix: WATER

Run Date: 06/28/2016
Run Time: 13:22
Analyst: BKT
Cal ID: ICP-MS2 - 28-JUN-16

Sample ID: WG574537-06
Method: 6020A
Units: ug/L

Analytes	MDL	RDL	Concentration	Qualifier
SILVER	.2	.8	.2	U
ARSENIC	.2	.8	.2	U
BARIUM	.6	2.4	.6	U
CADMIUM	.12	.48	.12	U
COBALT	.2	.8	.2	U
CHROMIUM	.4	1.6	.4	U
COPPER	.4	1.6	.4	U
MANGANESE	.4	1.6	.4	U
NICKEL	.8	3.2	.8	U
LEAD	.2	.8	.2	U
ANTIMONY	.2	.8	.405	*
THALLIUM	.04	.16	.04	U
VANADIUM	.2	.8	.2	U
ZINC	5	20	5	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4829718
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16060823
Run Date: 06/29/2016
Sample ID: WG574676-06
Instrument ID: ICP-MS2
Run Time: 11:41
Method: 6020A
File ID: NI.062916.114146
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS2 - 29-JUN-16
Matrix: WATER

Analytes	MDL	RDL	Concentration	Qualifier
SILVER	.2	.8	.2	U
ARSENIC	.2	.8	.2	U
BARIUM	.6	2.4	.6	U
CADMIUM	.12	.48	.12	U
COBALT	.2	.8	.2	U
CHROMIUM	.4	1.6	.4	U
COPPER	.4	1.6	.4	U
MANGANESE	.4	1.6	.4	U
NICKEL	.8	3.2	.8	U
LEAD	.2	.8	.2	U
ANTIMONY	.2	.8	.335	F
THALLIUM	.04	.16	.04	U
VANADIUM	.2	.8	.2	U
ZINC	5	20	5	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4829718
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-11
Instrument ID: ICP-MS2
Run Time: 13:38
Method: 6020A
File ID: NI.062816.133810
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.259	F
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829721
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-15
Instrument ID: ICP-MS2
Run Time: 14:34
Method: 6020A
File ID: NI.062816.143450
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829721
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-19
Instrument ID: ICP-MS2
Run Time: 15:13
Method: 6020A
File ID: NI.062816.151312
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.263	F
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829721
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-30
Instrument ID: ICP-MS2
Run Time: 18:52
Method: 6020A
File ID: NI.062816.185257
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829721
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-33
Instrument ID: ICP-MS2
Run Time: 20:26
Method: 6020A
File ID: NI.062816.202637
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829721
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-35
Instrument ID: ICP-MS2
Run Time: 21:04
Method: 6020A
File ID: NI.062816.210457
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829721
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/29/2016
Sample ID: WG574676-11
Instrument ID: ICP-MS2
Run Time: 12:07
Method: 6020A
File ID: NI.062916.120734
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829721
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/29/2016
Sample ID: WG574676-13
Instrument ID: ICP-MS2
Run Time: 12:40
Method: 6020A
File ID: NI.062916.124059
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4829721
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.

INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16060823
Run Date: 06/29/2016
Sample ID: WG574676-05
Instrument ID: ICP-MS2
Run Time: 11:38
Method: 6020A
File ID: NI.062916.113833
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Antimony		50	49.8	99.7	90 - 110	
Arsenic		50	50.9	102	90 - 110	
Barium		50	49.9	99.9	90 - 110	
Cadmium		50	50.3	101	90 - 110	
Chromium		50	50.0	100	90 - 110	
Cobalt		50	50.4	101	90 - 110	
Copper		50	50.8	102	90 - 110	
Lead		50	51.0	102	90 - 110	
Manganese		50	51.9	104	90 - 110	
Nickel		50	50.6	101	90 - 110	
Silver		50	50.9	102	90 - 110	
Thallium		50	50.7	101	90 - 110	
Vanadium		50	51.2	102	90 - 110	
Zinc		50	52.1	104	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4829717
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.

INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16060823 Run Date: 06/28/2016 Sample ID: WG574537-05
Instrument ID: ICP-MS2 Run Time: 13:18 Method: 6020A
File ID: NI.062816.131857 Analyst: BKT Units: ug/L
Workgroup (AAB#): WG574570 Cal ID: ICP-MS - 28-JUN-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Antimony		50	48.3	96.6	90 - 110	
Arsenic		50	50.3	101	90 - 110	
Barium		50	48.2	96.4	90 - 110	
Cadmium		50	49.0	98.0	90 - 110	
Chromium		50	50.0	100	90 - 110	
Cobalt		50	50.2	100	90 - 110	
Copper		50	50.1	100	90 - 110	
Lead		50	48.9	97.7	90 - 110	
Manganese		50	50.2	100	90 - 110	
Nickel		50	49.7	99.4	90 - 110	
Silver		50	49.3	98.6	90 - 110	
Thallium		50	49.1	98.2	90 - 110	
Vanadium		50	49.7	99.5	90 - 110	
Zinc		50	50.7	101	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4829717
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-10
Instrument ID: ICP-MS2
Run Time: 13:34
Method: 6020A
File ID: NI.062816.133459
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0492	mg/L	98.4	90 - 110		
Arsenic		0.0500	0.0495	mg/L	99.0	90 - 110		
Barium		0.0500	0.0487	mg/L	97.5	90 - 110		
Cadmium		0.0500	0.0487	mg/L	97.4	90 - 110		
Chromium		0.0500	0.0494	mg/L	98.8	90 - 110		
Cobalt		0.0500	0.0512	mg/L	102	90 - 110		
Copper		0.0500	0.0496	mg/L	99.2	90 - 110		
Lead		0.0500	0.0496	mg/L	99.3	90 - 110		
Manganese		0.0500	0.0497	mg/L	99.5	90 - 110		
Nickel		0.0500	0.0495	mg/L	98.9	90 - 110		
Silver		0.0500	0.0487	mg/L	97.3	90 - 110		
Thallium		0.0500	0.0489	mg/L	97.7	90 - 110		
Vanadium		0.0500	0.0499	mg/L	99.8	90 - 110		
Zinc		0.0500	0.0495	mg/L	99.0	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829720
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-14
Instrument ID: ICP-MS2
Run Time: 14:31
Method: 6020A
File ID: NI.062816.143138
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0480	mg/L	96.1	90 - 110		
Arsenic		0.0500	0.0501	mg/L	100	90 - 110		
Barium		0.0500	0.0474	mg/L	94.9	90 - 110		
Cadmium		0.0500	0.0480	mg/L	96.1	90 - 110		
Chromium		0.0500	0.0496	mg/L	99.1	90 - 110		
Cobalt		0.0500	0.0531	mg/L	106	90 - 110		
Copper		0.0500	0.0499	mg/L	99.7	90 - 110		
Lead		0.0500	0.0506	mg/L	101	90 - 110		
Manganese		0.0500	0.0492	mg/L	98.5	90 - 110		
Nickel		0.0500	0.0495	mg/L	99.0	90 - 110		
Silver		0.0500	0.0484	mg/L	96.8	90 - 110		
Thallium		0.0500	0.0490	mg/L	98.1	90 - 110		
Vanadium		0.0500	0.0508	mg/L	102	90 - 110		
Zinc		0.0500	0.0501	mg/L	100	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829720
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-18
Instrument ID: ICP-MS2
Run Time: 15:10
Method: 6020A
File ID: NI.062816.151001
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0504	mg/L	101	90 - 110		
Arsenic		0.0500	0.0499	mg/L	99.7	90 - 110		
Barium		0.0500	0.0485	mg/L	96.9	90 - 110		
Cadmium		0.0500	0.0487	mg/L	97.5	90 - 110		
Chromium		0.0500	0.0489	mg/L	97.9	90 - 110		
Cobalt		0.0500	0.0517	mg/L	103	90 - 110		
Copper		0.0500	0.0490	mg/L	97.9	90 - 110		
Lead		0.0500	0.0514	mg/L	103	90 - 110		
Manganese		0.0500	0.0487	mg/L	97.5	90 - 110		
Nickel		0.0500	0.0484	mg/L	96.8	90 - 110		
Silver		0.0500	0.0495	mg/L	98.9	90 - 110		
Thallium		0.0500	0.0506	mg/L	101	90 - 110		
Vanadium		0.0500	0.0498	mg/L	99.6	90 - 110		
Zinc		0.0500	0.0494	mg/L	98.9	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829720
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823 Run Date: 06/28/2016 Sample ID: WG574537-29
Instrument ID: ICP-MS2 Run Time: 18:49 Method: 6020A
File ID: NI.062816.184946 Analyst: BKT QC Key: DOD4
Workgroup (AAB#): WG574570 Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0481	mg/L	96.2	90 - 110		
Arsenic		0.0500	0.0471	mg/L	94.1	90 - 110		
Barium		0.0500	0.0464	mg/L	92.9	90 - 110		
Cadmium		0.0500	0.0463	mg/L	92.7	90 - 110		
Chromium		0.0500	0.0469	mg/L	93.8	90 - 110		
Cobalt		0.0500	0.0469	mg/L	93.8	90 - 110		
Copper		0.0500	0.0475	mg/L	94.9	90 - 110		
Lead		0.0500	0.0484	mg/L	96.7	90 - 110		
Manganese		0.0500	0.0473	mg/L	94.6	90 - 110		
Nickel		0.0500	0.0476	mg/L	95.1	90 - 110		
Silver		0.0500	0.0474	mg/L	94.8	90 - 110		
Thallium		0.0500	0.0474	mg/L	94.8	90 - 110		
Vanadium		0.0500	0.0472	mg/L	94.5	90 - 110		
Zinc		0.0500	0.0472	mg/L	94.5	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829720
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823 | Run Date: 06/28/2016 | Sample ID: WG574537-32
Instrument ID: ICP-MS2 | Run Time: 20:23 | Method: 6020A
File ID: NI.062816.202325 | Analyst: BKT | QC Key: DOD4
Workgroup (AAB#): WG574570 | Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0502	mg/L	100	90 - 110		
Arsenic		0.0500	0.0499	mg/L	99.8	90 - 110		
Barium		0.0500	0.0488	mg/L	97.7	90 - 110		
Cadmium		0.0500	0.0494	mg/L	98.8	90 - 110		
Chromium		0.0500	0.0508	mg/L	102	90 - 110		
Cobalt		0.0500	0.0517	mg/L	103	90 - 110		
Copper		0.0500	0.0507	mg/L	101	90 - 110		
Lead		0.0500	0.0515	mg/L	103	90 - 110		
Manganese		0.0500	0.0509	mg/L	102	90 - 110		
Nickel		0.0500	0.0501	mg/L	100	90 - 110		
Silver		0.0500	0.0500	mg/L	99.9	90 - 110		
Thallium		0.0500	0.0504	mg/L	101	90 - 110		
Vanadium		0.0500	0.0505	mg/L	101	90 - 110		
Zinc		0.0500	0.0506	mg/L	101	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829720
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823 Run Date: 06/28/2016 Sample ID: WG574537-34
Instrument ID: ICP-MS2 Run Time: 21:01 Method: 6020A
File ID: NI.062816.210146 Analyst: BKT QC Key: DOD4
Workgroup (AAB#): WG574570 Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0494	mg/L	98.8	90 - 110		
Arsenic		0.0500	0.0505	mg/L	101	90 - 110		
Barium		0.0500	0.0490	mg/L	98.1	90 - 110		
Cadmium		0.0500	0.0492	mg/L	98.4	90 - 110		
Chromium		0.0500	0.0506	mg/L	101	90 - 110		
Cobalt		0.0500	0.0511	mg/L	102	90 - 110		
Copper		0.0500	0.0509	mg/L	102	90 - 110		
Lead		0.0500	0.0534	mg/L	107	90 - 110		
Manganese		0.0500	0.0526	mg/L	105	90 - 110		
Nickel		0.0500	0.0509	mg/L	102	90 - 110		
Silver		0.0500	0.0485	mg/L	96.9	90 - 110		
Thallium		0.0500	0.0506	mg/L	101	90 - 110		
Vanadium		0.0500	0.0505	mg/L	101	90 - 110		
Zinc		0.0500	0.0506	mg/L	101	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829720
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/29/2016
Sample ID: WG574676-10
Instrument ID: ICP-MS2
Run Time: 12:00
Method: 6020A
File ID: NI.062916.120010
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0508	mg/L	102	90 - 110		
Arsenic		0.0500	0.0514	mg/L	103	90 - 110		
Barium		0.0500	0.0505	mg/L	101	90 - 110		
Cadmium		0.0500	0.0516	mg/L	103	90 - 110		
Chromium		0.0500	0.0504	mg/L	101	90 - 110		
Cobalt		0.0500	0.0538	mg/L	108	90 - 110		
Copper		0.0500	0.0507	mg/L	101	90 - 110		
Lead		0.0500	0.0529	mg/L	106	90 - 110		
Manganese		0.0500	0.0534	mg/L	107	90 - 110		
Nickel		0.0500	0.0502	mg/L	100	90 - 110		
Silver		0.0500	0.0508	mg/L	102	90 - 110		
Thallium		0.0500	0.0519	mg/L	104	90 - 110		
Vanadium		0.0500	0.0516	mg/L	103	90 - 110		
Zinc		0.0500	0.0510	mg/L	102	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829720
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/29/2016
Sample ID: WG574676-12
Instrument ID: ICP-MS2
Run Time: 12:37
Method: 6020A
File ID: NI.062916.123748
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0505	mg/L	101	90 - 110		
Arsenic		0.0500	0.0498	mg/L	99.6	90 - 110		
Barium		0.0500	0.0499	mg/L	99.8	90 - 110		
Cadmium		0.0500	0.0506	mg/L	101	90 - 110		
Chromium		0.0500	0.0489	mg/L	97.8	90 - 110		
Cobalt		0.0500	0.0506	mg/L	101	90 - 110		
Copper		0.0500	0.0500	mg/L	99.9	90 - 110		
Lead		0.0500	0.0499	mg/L	99.7	90 - 110		
Manganese		0.0500	0.0500	mg/L	99.9	90 - 110		
Nickel		0.0500	0.0491	mg/L	98.2	90 - 110		
Silver		0.0500	0.0505	mg/L	101	90 - 110		
Thallium		0.0500	0.0497	mg/L	99.4	90 - 110		
Vanadium		0.0500	0.0503	mg/L	101	90 - 110		
Zinc		0.0500	0.0498	mg/L	99.7	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4829720
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-07
Instrument ID: ICP-MS2
Run Time: 13:25
Method: 6020A
File ID: NI.062816.132523
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.414	ug/L	104	70 - 130		
Arsenic		0.400	0.419	ug/L	105	70 - 130		
Barium		0.750	0.681	ug/L	90.8	70 - 130		
Cadmium		0.240	0.234	ug/L	97.7	70 - 130		
Chromium		0.800	0.628	ug/L	78.5	70 - 130		
Cobalt		0.400	0.384	ug/L	96.1	70 - 130		
Copper		0.800	0.759	ug/L	94.9	70 - 130		
Lead		0.200	0.183	ug/L	91.6	70 - 130		
Manganese		0.500	0.437	ug/L	87.4	70 - 130		
Nickel		1.60	1.53	ug/L	95.5	70 - 130		
Silver		0.400	0.372	ug/L	92.9	70 - 130		
Thallium		0.0800	0.0768	ug/L	96.0	70 - 130		
Vanadium		0.400	0.332	ug/L	83.0	70 - 130		
Zinc		6.25	6.14	ug/L	98.3	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829722
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-20
Instrument ID: ICP-MS2
Run Time: 15:16
Method: 6020A
File ID: NI.062816.151629
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.429	ug/L	107	70 - 130		
Arsenic		0.400	0.377	ug/L	94.2	70 - 130		
Barium		0.750	0.677	ug/L	90.2	70 - 130		
Cadmium		0.240	0.235	ug/L	98.1	70 - 130		
Chromium		0.800	0.595	ug/L	74.4	70 - 130		
Cobalt		0.400	0.388	ug/L	97.1	70 - 130		
Copper		0.800	0.739	ug/L	92.4	70 - 130		
Lead		0.200	0.192	ug/L	96.0	70 - 130		
Manganese		0.500	0.408	ug/L	81.6	70 - 130		
Nickel		1.60	1.48	ug/L	92.6	70 - 130		
Silver		0.400	0.384	ug/L	96.0	70 - 130		
Thallium		0.0800	0.0804	ug/L	101	70 - 130		
Vanadium		0.400	0.327	ug/L	81.9	70 - 130		
Zinc		6.25	6.24	ug/L	99.9	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829722
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-31
Instrument ID: ICP-MS2
Run Time: 18:56
Method: 6020A
File ID: NI.062816.185610
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.386	ug/L	96.6	70 - 130		
Arsenic		0.400	0.381	ug/L	95.2	70 - 130		
Barium		0.750	0.671	ug/L	89.5	70 - 130		
Cadmium		0.240	0.217	ug/L	90.2	70 - 130		
Chromium		0.800	0.581	ug/L	72.7	70 - 130		
Cobalt		0.400	0.365	ug/L	91.2	70 - 130		
Copper		0.800	0.726	ug/L	90.7	70 - 130		
Lead		0.200	0.188	ug/L	93.9	70 - 130		
Manganese		0.500	0.396	ug/L	79.2	70 - 130		
Nickel		1.60	1.44	ug/L	89.7	70 - 130		
Silver		0.400	0.365	ug/L	91.2	70 - 130		
Thallium		0.0800	0.0752	ug/L	94.0	70 - 130		
Vanadium		0.400	0.309	ug/L	77.3	70 - 130		
Zinc		6.25	5.67	ug/L	90.8	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829722
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060823
Run Date: 06/28/2016
Sample ID: WG574537-44
Instrument ID: ICP-MS2
Run Time: 23:22
Method: 6020A
File ID: NI.062816.232221
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 28-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.384	ug/L	96.1	70 - 130		
Arsenic		0.400	0.418	ug/L	105	70 - 130		
Barium		0.750	0.696	ug/L	92.8	70 - 130		
Cadmium		0.240	0.223	ug/L	92.9	70 - 130		
Chromium		0.800	0.654	ug/L	81.8	70 - 130		
Cobalt		0.400	0.379	ug/L	94.6	70 - 130		
Copper		0.800	0.746	ug/L	93.2	70 - 130		
Lead		0.200	0.200	ug/L	100	70 - 130		
Manganese		0.500	0.444	ug/L	88.8	70 - 130		
Nickel		1.60	1.53	ug/L	95.6	70 - 130		
Silver		0.400	0.372	ug/L	93.0	70 - 130		
Thallium		0.0800	0.0774	ug/L	96.8	70 - 130		
Vanadium		0.400	0.335	ug/L	83.8	70 - 130		
Zinc		6.25	5.99	ug/L	95.8	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829722
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060823
Run Date: 06/29/2016
Sample ID: WG574676-07
Instrument ID: ICP-MS2
Run Time: 11:46
Method: 6020A
File ID: NI.062916.114653
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.456	ug/L	114	70 - 130		
Arsenic		0.400	0.429	ug/L	107	70 - 130		
Barium		0.750	0.733	ug/L	97.8	70 - 130		
Cadmium		0.240	0.237	ug/L	98.5	70 - 130		
Chromium		0.800	0.739	ug/L	92.4	70 - 130		
Cobalt		0.400	0.414	ug/L	103	70 - 130		
Copper		0.800	0.768	ug/L	96.0	70 - 130		
Lead		0.200	0.181	ug/L	90.3	70 - 130		
Manganese		0.500	0.468	ug/L	93.6	70 - 130		
Nickel		1.60	1.56	ug/L	97.7	70 - 130		
Silver		0.400	0.390	ug/L	97.5	70 - 130		
Thallium		0.0800	0.0871	ug/L	109	70 - 130		
Vanadium		0.400	0.382	ug/L	95.5	70 - 130		
Zinc		6.25	6.64	ug/L	106	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829722
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16060823
Run Date: 06/29/2016
Sample ID: WG574676-14
Instrument ID: ICP-MS2
Run Time: 12:44
Method: 6020A
File ID: NI.062916.124412
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574570
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.401	ug/L	100	70 - 130		
Arsenic		0.400	0.394	ug/L	98.6	70 - 130		
Barium		0.750	0.718	ug/L	95.7	70 - 130		
Cadmium		0.240	0.227	ug/L	94.7	70 - 130		
Chromium		0.800	0.768	ug/L	96.0	70 - 130		
Cobalt		0.400	0.405	ug/L	101	70 - 130		
Copper		0.800	0.751	ug/L	93.9	70 - 130		
Lead		0.200	0.179	ug/L	89.4	70 - 130		
Manganese		0.500	0.469	ug/L	93.8	70 - 130		
Nickel		1.60	1.53	ug/L	95.6	70 - 130		
Silver		0.400	0.383	ug/L	95.7	70 - 130		
Thallium		0.0800	0.0809	ug/L	101	70 - 130		
Vanadium		0.400	0.370	ug/L	92.4	70 - 130		
Zinc		6.25	6.89	ug/L	110	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4829722
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16060823
Instrument ID: ICP-MS2
Sol. A : WG574537-08 **File ID:** NI.062816.132834
Sol. AB : WG574537-09 **File ID:** NI.062816.133145

Workgroup (AAB#): WG574570
Method: 6020A
Units: ug/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Antimony	NS	0.0836	NS	100	94.7	94.7	
Arsenic	NS	-0.0118	NS	100	98.9	98.9	
Barium	NS	0.266	NS	100	94.5	94.5	
Cadmium	NS	-0.106	NS	100	95.6	95.6	
Chromium	NS	-0.0468	NS	100	96.5	96.5	
Cobalt	NS	0.0276	NS	100	97.7	97.7	
Copper	NS	0.231	NS	100	96.5	96.5	
Lead	NS	0.0352	NS	100	98.2	98.2	
Manganese	NS	0.0680	NS	100	97.7	97.7	
Nickel	NS	0.332	NS	100	96.0	96.0	
Silver	NS	0.0254	NS	100	89.5	89.5	
Thallium	NS	0.0257	NS	100	96.4	96.4	
Vanadium	NS	-0.0466	NS	100	96.5	96.5	
Zinc	NS	0.903	NS	100	101	101	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4829719
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16060823
Instrument ID: ICP-MS2
Sol. A : WG574537-16 **File ID:** NI.062816.150336
Sol. AB : WG574537-17 **File ID:** NI.062816.150647

Workgroup (AAB#): WG574570
Method: 6020A
Units: ug/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Antimony	NS	0.126	NS	100	101	101	
Arsenic	NS	-0.0170	NS	100	102	102	
Barium	NS	0.271	NS	100	98.3	98.3	
Cadmium	NS	-0.0956	NS	100	99.8	99.8	
Chromium	NS	-0.00590	NS	100	97.3	97.3	
Cobalt	NS	0.0242	NS	100	102	102	
Copper	NS	0.234	NS	100	97.4	97.4	
Lead	NS	0.0422	NS	100	102	102	
Manganese	NS	0.0683	NS	100	97.7	97.7	
Nickel	NS	0.343	NS	100	95.9	95.9	
Silver	NS	0.0228	NS	100	91.0	91.0	
Thallium	NS	0.0182	NS	100	99.2	99.2	
Vanadium	NS	-0.0922	NS	100	100	100	
Zinc	NS	0.968	NS	100	103	103	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4829719
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16060823
Instrument ID: ICP-MS2
Sol. A : WG574676-08 **File ID:** NI.062916.115053
Sol. AB : WG574676-09 **File ID:** NI.062916.115447

Workgroup (AAB#): WG574570
Method: 6020A
Units: ug/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Antimony	NS	0.0658	NS	100	100	100	
Arsenic	NS	0.0161	NS	100	105	105	
Barium	NS	0.303	NS	100	98.4	98.4	
Cadmium	NS	-0.103	NS	100	101	101	
Chromium	NS	0.0680	NS	100	101	101	
Cobalt	NS	0.0292	NS	100	105	105	
Copper	NS	0.245	NS	100	100	100	
Lead	NS	0.0174	NS	100	102	102	
Manganese	NS	0.0629	NS	100	105	105	
Nickel	NS	0.331	NS	100	100	100	
Silver	NS	0.0271	NS	100	95.3	95.3	
Thallium	NS	0.0271	NS	100	100	100	
Vanadium	NS	0.0706	NS	100	103	103	
Zinc	NS	1.14	NS	100	104	104	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4829719
Report generated 06/30/2016 09:24

Microbac Laboratories Inc.

INTERNAL STANDARD REPORT

Login: L16060823 **Analytical Method:** 6020
Analytical Workgroup: WG574570 **Matrix:** 1
Instrument: ICP-MS2 **Analyst:** BKT
ICAL Date: 28-JUN-2016 13:06

			BISMUTH	GERMANIUM	INDIUM
Sample	Type	Run Date	% Rec	% Rec	% Rec
L16060823-02	SAMP	28-JUN-2016 20:48	92.896	101.808	97.319
L16060823-03	SAMP	28-JUN-2016 20:52	93.551	101.87	99.333
L16060823-04	SAMP	28-JUN-2016 20:55	93.637	99.094	97.786
WG574242-02	BLANK	28-JUN-2016 19:51	99.476	99.537	97.857
WG574242-03	LCS	28-JUN-2016 19:54	99.956	99.501	98.669
WG574537-05	ICV	28-JUN-2016 13:18	106.21	104.263	104.758
WG574537-06	ICB	28-JUN-2016 13:22	103.902	104.951	103.59
WG574537-07	LLICV	28-JUN-2016 13:25	104.131	100.167	102.925
WG574537-08	ICS	28-JUN-2016 13:28	105.453	105.812	102.55
WG574537-09	ICS	28-JUN-2016 13:31	104.707	103.975	104.833
WG574537-10	CCV	28-JUN-2016 13:34	105.977	103.288	105.534
WG574537-11	CCB	28-JUN-2016 13:38	102.403	102.67	102.767
WG574537-14	CCV	28-JUN-2016 14:31	103.407	95.988	101.793
WG574537-15	CCB	28-JUN-2016 14:34	104.225	98.805	100.724
WG574537-16	ICS	28-JUN-2016 15:03	99.952	99.435	100.043
WG574537-17	ICS	28-JUN-2016 15:06	101.575	99.874	97.332
WG574537-18	CCV	28-JUN-2016 15:10	100.244	98.973	99.201
WG574537-19	CCB	28-JUN-2016 15:13	103.776	99.505	101.81
WG574537-20	LLCCV	28-JUN-2016 15:16	99.051	96.111	96.093
WG574537-29	CCV	28-JUN-2016 18:49	100.019	101.137	98.931
WG574537-30	CCB	28-JUN-2016 18:52	100.148	99.202	100.624
WG574537-31	LLCCV	28-JUN-2016 18:56	95.839	97.87	93.931
WG574537-32	CCV	28-JUN-2016 20:23	102.351	100.903	101.064
WG574537-33	CCB	28-JUN-2016 20:26	99.219	96.376	100.773
WG574537-34	CCV	28-JUN-2016 21:01	96.559	99.919	99.677
WG574537-35	CCB	28-JUN-2016 21:04	100.258	102.056	98.569
WG574537-44	LLCCV	28-JUN-2016 23:22	91.439	93.342	91.269
WG574570-01	PSPK	28-JUN-2016 20:13	100.708	100.256	100.218
WG574570-02	SERIAL	28-JUN-2016 20:17	103.535	103.697	101.2

Acceptance criteria: 30% - 120% Underlined recoveries are out of range
Acceptance criteria for CCVs and CCBs for method SW846-6020: 80% - 120%

INT_STD_ICPMS - Modified 07/28/2010
PDF File ID: 4829715
Report generated: 06/30/2016 09:24

Microbac Laboratories Inc.

INTERNAL STANDARD REPORT

Login: L16060823 **Analytical Method:** 6020
Analytical Workgroup: WG574570 **Matrix:** 1
Instrument: ICP-MS2 **Analyst:** BKT
ICAL Date: 29-JUN-2016 11:25

			BISMUTH	GERMANIUM	INDIUM
Sample	Type	Run Date	% Rec	% Rec	% Rec
L16060823-02	SAMP	29-JUN-2016 12:28	99.08	100.775	98.792
L16060823-03	SAMP	29-JUN-2016 12:31	99.415	102.35	99.462
L16060823-04	SAMP	29-JUN-2016 12:34	102.162	101.008	100.275
WG574676-05	ICV	29-JUN-2016 11:38	98.257	100.661	96.465
WG574676-06	ICB	29-JUN-2016 11:41	98.914	99.091	98.14
WG574676-07	LLICV	29-JUN-2016 11:46	95.79	95.394	93.877
WG574676-08	ICS	29-JUN-2016 11:50	98.596	98.274	96.165
WG574676-09	ICS	29-JUN-2016 11:54	97.891	98.362	95.991
WG574676-10	CCV	29-JUN-2016 12:00	96.559	97.448	96.281
WG574676-11	CCB	29-JUN-2016 12:07	98.969	97.431	96.855
WG574676-12	CCV	29-JUN-2016 12:37	99.303	102.045	96.376
WG574676-13	CCB	29-JUN-2016 12:40	100.991	102.01	98.548
WG574676-14	LLCCV	29-JUN-2016 12:44	96.535	95.69	95.451

Acceptance criteria: 30% - 120% Underlined recoveries are out of range
Acceptance criteria for CCVs and CCBs for method SW846-6020: 80% - 120%

INT_STD_ICPMS - Modified 07/28/2010
PDF File ID: 4829715
Report generated: 06/30/2016 09:24

Microbac Laboratories Inc.

INTERNAL STANDARD REPORT

Login: L16060823 **Analytical Method:** 6020
Analytical Workgroup: WG574570 **Matrix:** 2
Instrument: ICP-MS2 **Analyst:** BKT
ICAL Date: 28-JUN-2016 13:06

			BISMUTH	GERMANIUM	INDIUM
Sample	Type	Run Date	% Rec	% Rec	% Rec
L16060843-01	SAMP	28-JUN-2016 20:10	97.785	95.068	99.061

Acceptance criteria: 30% - 120% Underlined recoveries are out of range
Acceptance criteria for CCVs and CCBs for method SW846-6020: 80% - 120%

INT_STD_ICPMS - Modified 07/28/2010
PDF File ID: 4829715
Report generated: 06/30/2016 09:24

Microbac Laboratories Inc.
LINEAR RANGE (QUARTERLY)

Login Number: L16060823 **Date:** 04/19/2016
Insturment ID: ICP-MS2 **Method:** 6020A

Analyte	Integration Time (Sec.)	Concentration (ug/L)
Antimony	1.00	100.0
Arsenic	1.00	100.0
Barium	1.00	100.0
Cadmium	1.00	100.0
Chromium	1.00	100.0
Cobalt	1.00	100.0
Copper	1.00	100.0
Lead	1.00	100.0
Manganese	1.00	100.0
Nickel	1.00	100.0
Selenium	1.00	100.0
Silver	1.00	100.0
Thallium	1.00	100.0
Uranium	1.00	100.0
Vanadium	1.00	100.0
Zinc	1.00	100.0

Comments:
All analytes passed acceptance criteria at the specified concentration.

LINEAR_RANGE - Modified 03/06/2008
PDF File ID: 4829714
Report generated: 06/29/2016 08:34

2.3.2.3 Raw Data

MassCal File Name

Mass Calibration File Name Default.tun
MassCal File Path C:\NexIONData\MassCal\Default.tun
Peak Search Window: 1.00

Sample Information

Sample Date/Time: Tuesday, June 28, 2016 12:45:36

Mass Calibration and Resolution

Analyte	E Mass	Meas Mass	Mass C DAC Val	Res DAC Value	Meas Peak W	Custom Res
Li	7.016	7.025	1331	2024	0.706	
Mg	23.985	24.025	4508	2020	0.701	
Co	58.933	58.925	11686	2022	0.701	
In	114.904	114.875	22861	2028	0.696	
U	238.050	238.025	47455	2043	0.688	

Relative Std. Dev.

Mass	Meas. Intens. RSD
5.525	5.573
5.575	2.596
5.625	3.076
5.675	2.368
5.725	2.568
5.775	3.011
5.825	2.517
5.875	2.694
5.925	1.595
5.975	3.373
6.025	3.068
6.075	2.439
6.125	2.845
6.175	3.293
6.225	6.959
6.275	54.149
6.325	79.756
6.375	69.722
6.425	99.103
6.475	35.972
6.525	19.569
6.575	4.348
6.625	4.647
6.675	4.116
6.725	2.239
6.775	2.849
6.825	1.600

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 1

Approved: June 29, 2016

6.875	3.248
6.925	2.794
6.975	1.665
7.025	2.844
7.075	2.994
7.125	3.503
7.175	2.980
7.225	3.713
7.275	5.087
7.325	35.264
7.375	93.541
7.425	136.931
7.475	82.402
7.525	71.261
7.575	39.123
7.625	180.662
7.675	64.358
7.725	47.507
7.775	71.261
7.825	70.711
7.875	52.705
7.925	46.481
7.975	63.888
8.025	74.689
8.075	64.359
8.125	63.888
8.175	34.401
8.225	46.481
8.275	49.793
8.325	81.441
8.375	60.858
8.425	67.420
8.475	108.653
22.525	223.607
22.575	90.014
22.625	44.605
22.675	65.617
22.725	58.975
22.775	40.945
22.825	35.460
22.875	18.054
22.925	42.061
22.975	53.715
23.025	19.357
23.075	24.491
23.125	8.438
23.175	12.837

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 2

Approved: June 29, 2016

23.225	30.858
23.275	24.845
23.325	45.141
23.375	26.438
23.425	14.816
23.475	14.263
23.525	12.137
23.575	3.832
23.625	2.363
23.675	1.595
23.725	2.170
23.775	2.317
23.825	1.497
23.875	1.501
23.925	1.731
23.975	0.863
24.025	2.634
24.075	2.734
24.125	2.053
24.175	2.398
24.225	2.993
24.275	2.883
24.325	7.686
24.375	28.384
24.425	36.617
24.475	12.141
24.525	3.189
24.575	0.926
24.625	1.585
24.675	1.547
24.725	2.051
24.775	1.408
24.825	2.678
24.875	1.711
24.925	1.949
24.975	2.595
25.025	3.005
25.075	2.962
25.125	2.366
25.175	2.463
25.225	2.722
25.275	9.971
25.325	31.599
25.375	47.140
25.425	64.885
25.475	6.734
57.525	8.866

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 3

Approved: June 29, 2016

57.575	5.291
57.625	3.440
57.675	3.697
57.725	3.076
57.775	2.567
57.825	2.416
57.875	1.794
57.925	2.766
57.975	2.199
58.025	1.926
58.075	2.254
58.125	1.436
58.175	1.390
58.225	3.504
58.275	5.534
58.325	34.267
58.375	33.055
58.425	20.772
58.475	14.356
58.525	5.854
58.575	3.115
58.625	4.057
58.675	2.315
58.725	2.604
58.775	2.171
58.825	1.290
58.875	1.298
58.925	2.935
58.975	1.929
59.025	1.078
59.075	1.921
59.125	2.163
59.175	2.388
59.225	3.289
59.275	8.929
59.325	38.030
59.375	58.736
59.425	28.284
59.475	17.563
59.525	12.128
59.575	6.526
59.625	2.279
59.675	2.377
59.725	6.247
59.775	4.326
59.825	2.635
59.875	3.246

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 4

Approved: June 29, 2016

59.925	1.954
59.975	3.753
60.025	5.882
60.075	3.380
60.125	2.788
60.175	3.075
60.225	4.430
60.275	18.947
60.325	43.241
60.375	22.017
60.425	79.756
60.475	50.461
113.525	7.111
113.575	1.614
113.625	3.295
113.675	3.378
113.725	2.143
113.775	2.366
113.825	3.191
113.875	4.581
113.925	1.498
113.975	2.463
114.025	3.675
114.075	2.767
114.125	4.064
114.175	2.276
114.225	6.994
114.275	28.008
114.325	41.691
114.375	17.526
114.425	9.127
114.475	6.547
114.525	3.985
114.575	2.757
114.625	3.785
114.675	2.945
114.725	3.014
114.775	2.252
114.825	1.319
114.875	1.984
114.925	1.415
114.975	0.577
115.025	0.422
115.075	1.851
115.125	0.695
115.175	3.426
115.225	3.732

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 5

Approved: June 29, 2016

115.275	9.818
115.325	57.810
115.375	52.705
115.425	23.458
115.475	18.657
115.525	4.767
115.575	6.521
115.625	4.984
115.675	3.649
115.725	7.463
115.775	7.302
115.825	2.294
115.875	3.587
115.925	2.925
115.975	3.587
116.025	4.548
116.075	6.187
116.125	2.246
116.175	11.560
116.225	3.389
116.275	17.538
116.325	51.618
116.375	52.705
116.425	42.127
116.475	34.312
236.525	
236.575	28.853
236.625	56.165
236.675	51.701
236.725	11.620
236.775	50.775
236.825	27.217
236.875	33.431
236.925	29.881
236.975	45.955
237.025	46.771
237.075	7.402
237.125	30.162
237.175	24.686
237.225	34.297
237.275	16.110
237.325	20.328
237.375	28.770
237.425	26.702
237.475	18.887
237.525	24.816
237.575	6.725

Report Date/Time: Tuesday, June 28, 2016 12:48:47
Page 6

Approved: June 29, 2016

237.625	8.740
237.675	3.668
237.725	2.023
237.775	1.587
237.825	3.110
237.875	1.530
237.925	2.794
237.975	3.068
238.025	2.967
238.075	2.352
238.125	2.652
238.175	2.474
238.225	2.129
238.275	1.500
238.325	2.197
238.375	3.421
238.425	3.111
238.475	6.669
238.525	16.077
238.575	17.557
238.625	21.969
238.675	32.775
238.725	22.232
238.775	26.146
238.825	21.694
238.875	19.088
238.925	8.384
238.975	17.180
239.025	59.843
239.075	31.485
239.125	33.497
239.175	31.181
239.225	27.834
239.275	28.743
239.325	26.503
239.375	17.174
239.425	9.195
239.475	35.446

Approved: June 29, 2016

SmartTune Wizard - Summary

```
Optimization Summary

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Start Time: 6/28/2016 12:50:43 PM
End Time: 6/28/2016 12:52:55 PM

Daily Performance Check - [Passed]  Optimum value(s): N/A
        Obtained Intensity (Be 9.0122): 5221.30
        Obtained Intensity (Mg 23.985): 136183.87
        Obtained Intensity (In 114.904): 66441.11
        Obtained Intensity (U 238.05): 147068.09
        Obtained Intensity (Bkgd 220): 0.20
        Obtained Formula (CeO 155.9 / Ce 139.905): 0.018 (=3731.53 / 205818.69)
        Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.004 (=890.23 / 205818.69)
```

Report Date/Time: Tuesday, June 28, 2016 12:52:56
Page 1

Approved: June 29, 2016

SmartTune Wizard - Details

```
Optimization Details

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Optimization Status

Start Time: 6/28/2016 12:50:43 PM

Daily Performance Check
        Optimization Settings:
                Method: C:\NexIONData\Method\ESI Daily Performance.mth.
                Intensity Criterion: Be 9.0122 > 2000
                Intensity Criterion: Mg 23.985 > 15000
                Intensity Criterion: In 114.904 > 40000
                Intensity Criterion: U 238.05 > 30000
                Intensity Criterion: Bkgd 220 <= 5
                Formula Criterion: CeO 155.9 / Ce 139.905 <= 0.025
                Formula Criterion: Ce++ 69.9527 / Ce 139.905 <= 0.03

        Optimization Results:
        Initial Try
                Obtained Intensity (Be 9.0122): 5221.30
                Obtained Intensity (Mg 23.985): 136183.87
                Obtained Intensity (In 114.904): 66441.11
                Obtained Intensity (U 238.05): 147068.09
                Obtained Intensity (Bkgd 220): 0.20
                Obtained Formula (CeO 155.9 / Ce 139.905): 0.018 (=3731.53 / 205818.69)
                Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.004 (=890.23 / 205818.69)
[Passed]  Optimum value(s): N/A

End Time: 6/28/2016 12:52:55 PM
```


Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: Blank
Sample Date/Time: Tuesday, June 28, 2016 13:02:57
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	81760.1	2.0				ug/L		Standard
	Be	9	16.7	75.5				ug/L		Standard
	Al	27	1410.1	2.3				ug/L		Standard
	Sc	45	39543.2	2.7				ug/L		Standard
	Ti	47	39.0	9.2				ug/L		Standard
	V	51	1415.2	6.8				ug/L		Standard
	Cr	52	6946.3	1.7				ug/L		Standard
	Cr	53	328.3	14.1				ug/L		Standard
	Mn	55	823.0	8.8				ug/L		Standard
	Co	59	251.7	9.6				ug/L		Standard
	Ni	60	192.0	8.0				ug/L		Standard
	Cu	65	140.7	5.3				ug/L		Standard
	Zn	66	490.7	3.4				ug/L		Standard
>	Ge	72	515086.9	1.2				ug/L		Standard
	As	75	19.8	169.9				ug/L		Standard
	Se	82	15.3	19.8				ug/L		Standard
	Se-1	77	48.0	15.7				ug/L		Standard
>	Ga	71	90.0	5.6				mg/L		Standard
	Rb	85	23.3	68.9				ug/L		Standard
	Y	89	448509.9	0.5				ug/L		Standard
>	Rh	103	15.0	66.7				ug/L		Standard
	Mo	98	10.9	19.2				ug/L		Standard
	Ag	107	134.0	44.7				ug/L		Standard
	Cd	111	6.6	57.0				mg/L		Standard
	Cd	114	16.1	68.4				ug/L		Standard
>	In	115	611116.5	1.9				ug/L		Standard
	Sn	118	518.3	7.2				ug/L		Standard
	Sb	123	126.7	13.8				ug/L		Standard
	Ba	135	45.3	39.8				ug/L		Standard
	Ce	140	298.3	72.3				ug/L		Standard
>	Tb	159	977942.3	1.0				ug/L		Standard
	Ho	165	11.7	24.7				ug/L		Standard
	Tl	203	30.7	114.0				ug/L		Standard
	Tl	205	20.0	152.1				ug/L		Standard
	Pb	206	285.7	7.2				ug/L		Standard
	Pb	207	257.0	5.7				ug/L		Standard
	Pb	208	955.7	3.9				ug/L		Standard
	U	238	17.0	67.8				ug/L		Standard
>	Bi	209	557080.4	0.4				ug/L		Standard

Sample ID: Blank
Report Date/Time: Tuesday, June 28, 2016 13:05:14
Page 1

Approved: June 29, 2016

	Na	23	0.0		mg/L	Standard
	Mg	24	31.7	36.5	mg/L	Standard
	K	39	30.0	57.7	mg/L	Standard
	Ca	43	48.3	43.1	mg/L	Standard
	Fe	54	159.3	19.3	mg/L	Standard
	Fe	57	195.0	11.2	mg/L	Standard
>	**Sc-1**	45	39543.2	2.7	mg/L	Standard
	Cl	35	27641.9	2.8	ug/L	Standard
	Kr	83	2.0	50.0	ug/L	Standard
	Br	81	1646.8	11.1	ug/L	Standard
	P	31	15231.0	1.6	ug/L	Standard
	S	34	1845.1	7.5	ug/L	Standard
	Sr	88	70.0	12.4	ug/L	Standard
	C	12	113.3	28.4	mg/L	Standard
	N	14	0.0		mg/L	Standard
	Hg	202	0.0		mg/L	Standard
	Dy	164	12.4	122.2	mg/L	Standard
	Ho-1	165	11.7	24.7	mg/L	Standard
	Er	166	20.0	50.0	mg/L	Standard
	I	127	1161.7	12.0	mg/L	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Blank
Report Date/Time: Tuesday, June 28, 2016 13:05:14
Page 2

Approved: June 29, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Blank
Report Date/Time: Tuesday, June 28, 2016 13:05:14
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: Standard 1
Sample Date/Time: Tuesday, June 28, 2016 13:06:09
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83638.6	0.9				ug/L	81760	Standard
	Be	9	11.7	49.5				ug/L	17	Standard
	Al	27	1445.1	4.5				ug/L	1410	Standard
	Sc	45	40184.9	2.0				ug/L	39543	Standard
	Ti	47	38.3	12.0				ug/L	39	Standard
	V	51	1492.6	2.6				ug/L	1415	Standard
	Cr	52	7409.5	2.1				ug/L	6946	Standard
	Cr	53	321.7	15.3				ug/L	328	Standard
	Mn	55	835.0	1.8				ug/L	823	Standard
	Co	59	226.7	11.0				ug/L	252	Standard
	Ni	60	182.7	5.1				ug/L	192	Standard
	Cu	65	134.7	15.9				ug/L	141	Standard
	Zn	66	519.7	4.6				ug/L	491	Standard
>	**Ge**	72	517782.5	2.4				ug/L	515087	Standard
	As	75	-33.8	72.0				ug/L	20	Standard
	Se	82	16.9	27.2				ug/L	15	Standard
	Se-1	77	48.7	14.9				ug/L	48	Standard
>	**Ga**	71	95.0	19.0				mg/L	90	Standard
	Rb	85	33.3	82.6				ug/L	23	Standard
	Y	89	456301.2	1.1				ug/L	448510	Standard
>	**Rh**	103	21.7	58.1				ug/L	15	Standard
	Mo	98	8.3	58.9				ug/L	11	Standard
	Ag	107	79.0	9.1				ug/L	134	Standard
	Cd	111	3.3	17.5				mg/L	7	Standard
	Cd	114	6.2	12.1				ug/L	16	Standard
>	**In**	115	624714.7	2.4				ug/L	611117	Standard
	Sn	118	515.0	5.4				ug/L	518	Standard
	Sb	123	47.7	15.0				ug/L	127	Standard
	Ba	135	33.7	27.6				ug/L	45	Standard
	Ce	140	86.7	24.0				ug/L	298	Standard
>	**Tb**	159	1001347.1	1.8				ug/L	977942	Standard
	Ho	165	6.7	43.3				ug/L	12	Standard
	Tl	203	5.3	60.3				ug/L	31	Standard
	Tl	205	3.3	86.6				ug/L	20	Standard
	Pb	206	276.7	3.7				ug/L	286	Standard
	Pb	207	213.3	7.3				ug/L	257	Standard
	Pb	208	933.0	3.4				ug/L	956	Standard
	U	238	4.7	12.4				ug/L	17	Standard
>	**Bi**	209	562888.8	0.9				ug/L	557080	Standard

Sample ID: Standard 1
Report Date/Time: Tuesday, June 28, 2016 13:08:26
Page 1

Approved: June 29, 2016

	Na	23	0.0		mg/L	0	Standard
	Mg	24	36.7	7.9	mg/L	32	Standard
	K	39	28.3	44.4	mg/L	30	Standard
	Ca	43	36.7	28.4	mg/L	48	Standard
	Fe	54	172.5	32.0	mg/L	159	Standard
	Fe	57	211.7	34.4	mg/L	195	Standard
>	**Sc-1**	45	40184.9	2.0	mg/L	39543	Standard
	Cl	35	29114.6	1.1	ug/L	27642	Standard
	Kr	83	3.3	96.4	ug/L	2	Standard
	Br	81	1676.8	6.8	ug/L	1647	Standard
	P	31	15918.4	1.2	ug/L	15231	Standard
	S	34	1885.1	11.3	ug/L	1845	Standard
	Sr	88	58.3	21.6	ug/L	70	Standard
	C	12	116.7	24.7	mg/L	113	Standard
	N	14	0.0		mg/L	0	Standard
	Hg	202	0.0		mg/L	0	Standard
	Dy	164	16.2	93.4	mg/L	12	Standard
	Ho-1	165	6.7	43.3	mg/L	12	Standard
	Er	166	10.0	100.0	mg/L	20	Standard
	I	127	1170.0	16.3	mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Standard 1
Report Date/Time: Tuesday, June 28, 2016 13:08:26
Page 2

Approved: June 29, 2016

⎣	Rb	85
⎡	Y	89
⎣>	Rh	103
⎡	Mo	98
⎢	Ag	107
⎢	Cd	111
⎢	Cd	114
⎢>	In	115
⎢	Sn	118
⎢	Sb	123
⎣	Ba	135
⎡	Ce	140
⎣>	Tb	159
⎡	Ho	165
⎢	Tl	203
⎢	Tl	205
⎢	Pb	206
⎢	Pb	207
⎢	Pb	208
⎢	U	238
⎣>	Bi	209
⎡	Na	23
⎢	Mg	24
⎢	K	39
⎢	Ca	43
⎢	Fe	54
⎢	Fe	57
⎣>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 1
Report Date/Time: Tuesday, June 28, 2016 13:08:26
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: Standard 2
Sample Date/Time: Tuesday, June 28, 2016 13:09:21
Number of Replicates: 3
Autosampler Position: 2
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82626.5	2.9				ug/L	81760	Standard
	Be	9	60.0	14.4				ug/L	17	Standard
	Al	27	10023.3	4.0				ug/L	1410	Standard
	Sc	45	41007.1	1.4				ug/L	39543	Standard
	Ti	47	69.0	5.8				ug/L	39	Standard
	V	51	1868.4	2.4				ug/L	1415	Standard
	Cr	52	8071.5	1.1				ug/L	6946	Standard
	Cr	53	360.0	4.2				ug/L	328	Standard
	Mn	55	1347.7	1.9				ug/L	823	Standard
	Co	59	599.0	5.9				ug/L	252	Standard
	Ni	60	305.7	5.8				ug/L	192	Standard
	Cu	65	280.3	2.3				ug/L	141	Standard
	Zn	66	1008.0	4.7				ug/L	491	Standard
>	**Ge**	72	539572.1	1.8				ug/L	515087	Standard
	As	75	11.1	230.1				ug/L	20	Standard
	Se	82	20.4	17.1				ug/L	15	Standard
	Se-1	77	60.7	1.9				ug/L	48	Standard
>	**Ga**	71	83.3	19.3				mg/L	90	Standard
	Rb	85	31.7	48.2				ug/L	23	Standard
	Y	89	463986.2	1.2				ug/L	448510	Standard
>	**Rh**	103	13.3	57.3				ug/L	15	Standard
	Mo	98	329.5	2.8				ug/L	11	Standard
	Ag	107	455.0	1.5				ug/L	134	Standard
	Cd	111	115.7	6.5				mg/L	7	Standard
	Cd	114	292.0	6.0				ug/L	16	Standard
>	**In**	115	625608.8	0.7				ug/L	611117	Standard
	Sn	118	920.0	9.2				ug/L	518	Standard
	Sb	123	303.4	7.6				ug/L	127	Standard
	Ba	135	158.7	8.0				ug/L	45	Standard
	Ce	140	88.3	42.5				ug/L	298	Standard
>	**Tb**	159	1002874.0	1.4				ug/L	977942	Standard
	Ho	165	11.7	24.7				ug/L	12	Standard
	Tl	203	472.7	2.0				ug/L	31	Standard
	Tl	205	391.7	11.0				ug/L	20	Standard
	Pb	206	628.7	5.7				ug/L	286	Standard
	Pb	207	539.3	6.7				ug/L	257	Standard
	Pb	208	2133.7	2.7				ug/L	956	Standard
	U	238	388.7	5.1				ug/L	17	Standard
>	**Bi**	209	570508.1	0.4				ug/L	557080	Standard

Sample ID: Standard 2
Report Date/Time: Tuesday, June 28, 2016 13:11:38
Page 1

Approved: June 29, 2016

	Na	23	0.0		mg/L	0	Standard
	Mg	24	48.3	26.0	mg/L	32	Standard
	K	39	45.0	19.2	mg/L	30	Standard
	Ca	43	36.7	43.8	mg/L	48	Standard
	Fe	54	162.5	17.6	mg/L	159	Standard
	Fe	57	193.3	3.0	mg/L	195	Standard
>	**Sc-1**	45	41007.1	1.4	mg/L	39543	Standard
	Cl	35	29715.1	0.6	ug/L	27642	Standard
	Kr	83	3.0	33.3	ug/L	2	Standard
	Br	81	1836.8	4.6	ug/L	1647	Standard
	P	31	16030.1	3.1	ug/L	15231	Standard
	S	34	1895.1	1.4	ug/L	1845	Standard
	Sr	88	61.7	9.4	ug/L	70	Standard
	C	12	143.3	21.3	mg/L	113	Standard
	N	14	0.0		mg/L	0	Standard
	Hg	202	0.0		mg/L	0	Standard
	Dy	164	19.5	50.0	mg/L	12	Standard
	Ho-1	165	11.7	24.7	mg/L	12	Standard
	Er	166	10.0	100.0	mg/L	20	Standard
	I	127	1208.4	23.1	mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Standard 2
Report Date/Time: Tuesday, June 28, 2016 13:11:38
Page 2

Approved: June 29, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 2
Report Date/Time: Tuesday, June 28, 2016 13:11:38
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: Standard 3
Sample Date/Time: Tuesday, June 28, 2016 13:12:32
Number of Replicates: 3
Autosampler Position: 3
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83151.0	2.6				ug/L	81760	Standard
	Be	9	50594.4	3.3	**50.0000**	0.655	1.3	ug/L	17	Standard
	Al	27	7832644.8	2.3	**50.0000**	1.818	3.6	ug/L	1410	Standard
	Sc	45	42972.5	0.7				ug/L	39543	Standard
	Ti	47	26044.7	0.2	**100.0000**	1.044	1.0	ug/L	39	Standard
	V	51	369925.9	1.5	**50.0000**	0.431	0.9	ug/L	1415	Standard
	Cr	52	376035.2	0.6	**50.0000**	0.381	0.8	ug/L	6946	Standard
	Cr	53	47201.8	2.5	**50.0000**	1.648	3.3	ug/L	328	Standard
	Mn	55	339840.1	0.9	**50.0000**	0.028	0.1	ug/L	823	Standard
	Co	59	369835.8	1.4	**50.0000**	0.329	0.7	ug/L	252	Standard
	Ni	60	100390.8	0.3	**50.0000**	0.410	0.8	ug/L	192	Standard
	Cu	65	101743.1	0.9	**50.0000**	0.203	0.4	ug/L	141	Standard
	Zn	66	59304.5	0.9	**50.0000**	0.371	0.7	ug/L	491	Standard
>	**Ge**	72	547553.1	0.8				ug/L	515087	Standard
	As	75	59263.2	0.8	**50.0000**	0.205	0.4	ug/L	20	Standard
	Se	82	6019.0	1.0	**50.0000**	0.596	1.2	ug/L	15	Standard
	Se-1	77	4068.9	1.3	**50.0000**	0.434	0.9	ug/L	48	Standard
>	**Ga**	71	120.0	22.0				mg/L	90	Standard
	Rb	85	803.4	2.6				ug/L	23	Standard
	Y	89	462085.4	0.0				ug/L	448510	Standard
>	**Rh**	103	33.3	52.7				ug/L	15	Standard
	Mo	98	278482.4	0.5	**100.0000**	0.941	0.9	ug/L	11	Standard
	Ag	107	351728.0	0.3	**50.0000**	0.642	1.3	ug/L	134	Standard
	Cd	111	110396.3	1.4	**50.0000**	0.556	1.1	mg/L	7	Standard
	Cd	114	267604.2	1.4	**50.0000**	0.663	1.3	ug/L	16	Standard
>	**In**	115	627830.2	1.2				ug/L	611117	Standard
	Sn	118	304675.0	1.7	**50.0000**	0.618	1.2	ug/L	518	Standard
	Sb	123	244739.5	0.4	**50.0000**	0.675	1.4	ug/L	127	Standard
	Ba	135	106555.1	1.0	**50.0000**	0.097	0.2	ug/L	45	Standard
	Ce	140	116.7	9.9				ug/L	298	Standard
>	**Tb**	159	1018565.7	2.1				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	438004.0	1.2	**50.0000**	0.427	0.9	ug/L	31	Standard
	Tl	205	379940.6	1.7	**50.0000**	0.597	1.2	ug/L	20	Standard
	Pb	206	284208.1	0.3	**50.0000**	0.335	0.7	ug/L	286	Standard
	Pb	207	256392.2	0.5	**50.0000**	0.374	0.7	ug/L	257	Standard
	Pb	208	978425.0	1.4	**50.0000**	0.588	1.2	ug/L	956	Standard
	U	238	371095.6	0.5	**50.0000**	0.490	1.0	ug/L	17	Standard
>	**Bi**	209	583572.7	0.6				ug/L	557080	Standard

Sample ID: Standard 3
Report Date/Time: Tuesday, June 28, 2016 13:14:49
Page 1

Approved: June 29, 2016

Na	23	0.0					mg/L	0	Standard
Mg	24	4002.2	2.2	**5.0000**	0.144	2.9	mg/L	32	Standard
K	39	1640.1	4.4	**5.0000**	0.220	4.4	mg/L	30	Standard
Ca	43	73.3	27.6	**5.0000**	2.832	56.6	mg/L	48	Standard
Fe	54	6006.1	1.4	**5.0000**	0.034	0.7	mg/L	159	Standard
Fe	57	1818.4	10.7	**5.0000**	0.606	12.1	mg/L	195	Standard
Sc-1	45	42972.5	0.7				mg/L	39543	Standard
Cl	35	31584.9	0.4				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1736.8	12.0				ug/L	1647	Standard
P	31	18696.5	1.2				ug/L	15231	Standard
S	34	2423.5	4.1				ug/L	1845	Standard
Sr	88	71.7	21.3				ug/L	70	Standard
C	12	110.0	39.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	18.6	110.3				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	30.0	33.3				mg/L	20	Standard
I	127	1405.1	8.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72			
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: Standard 3
Report Date/Time: Tuesday, June 28, 2016 13:14:49
Page 2

Approved: June 29, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: Standard 4
Sample Date/Time: Tuesday, June 28, 2016 13:15:44
Number of Replicates: 3
Autosampler Position: 4
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81919.2	1.8				ug/L	81760	Standard
	Be	9	99366.8	4.2	**99.8402**	3.215	3.2	ug/L	17	Standard
	Al	27	15456145.8	2.8	**100.0514**	1.254	1.3	ug/L	1410	Standard
	Sc	45	41889.5	2.4				ug/L	39543	Standard
	Ti	47	52065.7	3.9	**203.2707**	7.861	3.9	ug/L	39	Standard
	V	51	718696.0	1.3	**100.2818**	2.594	2.6	ug/L	1415	Standard
	Cr	52	727396.2	2.7	**100.4747**	2.418	2.4	ug/L	6946	Standard
	Cr	53	90202.9	3.4	**99.4979**	3.023	3.0	ug/L	328	Standard
	Mn	55	679268.3	1.6	**101.6708**	2.363	2.3	ug/L	823	Standard
	Co	59	730686.9	1.6	**101.0312**	2.535	2.5	ug/L	252	Standard
	Ni	60	193870.4	2.0	**99.9159**	1.756	1.8	ug/L	192	Standard
	Cu	65	194619.8	1.4	**99.4389**	1.906	1.9	ug/L	141	Standard
	Zn	66	113576.2	1.7	**99.8564**	1.521	1.5	ug/L	491	Standard
>	**Ge**	72	530148.8	1.5				ug/L	515087	Standard
	As	75	116520.9	0.8	**100.7477**	1.604	1.6	ug/L	20	Standard
	Se	82	12083.5	1.4	**101.8724**	2.244	2.2	ug/L	15	Standard
	Se-1	77	7795.3	2.5	**99.8242**	2.540	2.5	ug/L	48	Standard
>	**Ga**	71	125.0	22.3				mg/L	90	Standard
	Rb	85	1618.4	10.1				ug/L	23	Standard
	Y	89	461192.2	0.4				ug/L	448510	Standard
>	**Rh**	103	56.7	20.4				ug/L	15	Standard
	Mo	98	556218.8	1.0	**199.4538**	3.118	1.6	ug/L	11	Standard
	Ag	107	678834.5	1.8	**98.0160**	2.327	2.4	ug/L	134	Standard
	Cd	111	219330.5	1.3	**99.4613**	1.808	1.8	mg/L	7	Standard
	Cd	114	534582.7	2.1	**99.7362**	2.615	2.6	ug/L	16	Standard
>	**In**	115	630478.6	0.6				ug/L	611117	Standard
	Sn	118	609964.7	1.0	**99.8933**	1.520	1.5	ug/L	518	Standard
	Sb	123	482106.0	1.2	**99.0353**	1.710	1.7	ug/L	127	Standard
	Ba	135	212119.9	0.9	**99.5711**	1.444	1.4	ug/L	45	Standard
	Ce	140	165.0	42.7				ug/L	298	Standard
>	**Tb**	159	1009801.9	0.4				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	865813.4	1.1	**100.7605**	1.348	1.3	ug/L	31	Standard
	Tl	205	752692.6	1.1	**100.8681**	0.657	0.7	ug/L	20	Standard
	Pb	206	549987.8	0.2	**99.7256**	0.843	0.8	ug/L	286	Standard
	Pb	207	495989.2	0.5	**99.7047**	0.496	0.5	ug/L	257	Standard
	Pb	208	1945321.0	0.5	**101.0779**	0.484	0.5	ug/L	956	Standard
	U	238	738847.3	0.4	**101.1182**	1.306	1.3	ug/L	17	Standard
>	**Bi**	209	568134.9	1.0				ug/L	557080	Standard

Sample ID: Standard 4
Report Date/Time: Tuesday, June 28, 2016 13:18:01
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**19.9850**	17.308	86.6	mg/L	0	Standard
Mg	24	8245.6	0.3	**10.3081**	0.247	2.4	mg/L	32	Standard
K	39	3202.0	9.6	**10.0847**	1.159	11.5	mg/L	30	Standard
Ca	43	108.3	26.2	**10.2160**	4.199	41.1	mg/L	48	Standard
Fe	54	12180.7	3.3	**10.2708**	0.331	3.2	mg/L	159	Standard
Fe	57	3313.7	6.4	**9.9500**	0.928	9.3	mg/L	195	Standard
Sc-1	45	41889.5	2.4				mg/L	39543	Standard
Cl	35	30969.0	3.8				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1773.4	7.0				ug/L	1647	Standard
P	31	19072.0	2.1				ug/L	15231	Standard
S	34	2495.2	5.2				ug/L	1845	Standard
Sr	88	56.7	5.1				ug/L	70	Standard
C	12	143.3	29.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	13.0	113.4				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	6.7	173.2				mg/L	20	Standard
I	127	1551.7	12.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72			
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: Standard 4
Report Date/Time: Tuesday, June 28, 2016 13:18:01
Page 2

Approved: June 29, 2016

	Rb	85
	Y	89
>	Rh	103
	Mo	98
	Ag	107
	Cd	111
	Cd	114
>	In	115
	Sn	118
	Sb	123
	Ba	135
	Ce	140
>	Tb	159
	Ho	165
	Tl	203
	Tl	205
	Pb	206
	Pb	207
	Pb	208
	U	238
>	Bi	209
	Na	23
	Mg	24
	K	39
	Ca	43
	Fe	54
	Fe	57
>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Corr. Coef.	Na	23	Correlation coefficient < 0.998

Sample ID: Standard 4
Report Date/Time: Tuesday, June 28, 2016 13:18:01
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 1
Sample Date/Time: Tuesday, June 28, 2016 13:18:57
Number of Replicates: 3
Autosampler Position: 201
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81910.8	1.2				ug/L	81760	Standard
	Be	9	50188.0	2.5	**50.4285**	0.671	1.3	ug/L	17	Standard
	Al	27	7919939.7	1.3	**51.2716**	0.184	0.4	ug/L	1410	Standard
	Sc	45	42151.9	1.0				ug/L	39543	Standard
	Ti	47	26799.3	0.9	**103.2216**	1.931	1.9	ug/L	39	Standard
	V	51	362002.1	2.3	**49.7477**	0.555	1.1	ug/L	1415	Standard
	Cr	52	370323.8	0.7	**49.9789**	0.741	1.5	ug/L	6946	Standard
	Cr	53	45478.3	5.4	**49.3268**	1.623	3.3	ug/L	328	Standard
	Mn	55	340081.6	0.1	**50.1792**	1.001	2.0	ug/L	823	Standard
	Co	59	367567.5	0.6	**50.1558**	0.780	1.6	ug/L	252	Standard
	Ni	60	97767.5	1.8	**49.6873**	0.150	0.3	ug/L	192	Standard
	Cu	65	99441.7	2.0	**50.1064**	0.391	0.8	ug/L	141	Standard
	Zn	66	58882.6	2.1	**50.6983**	0.019	0.0	ug/L	491	Standard
>	**Ge**	72	537043.0	2.1				ug/L	515087	Standard
	As	75	58926.0	1.4	**50.3136**	0.381	0.8	ug/L	20	Standard
	Se	82	6003.9	1.4	**49.9034**	0.396	0.8	ug/L	15	Standard
	Se-1	77	4037.5	1.3	**50.7056**	1.619	3.2	ug/L	48	Standard
>	**Ga**	71	113.3	22.2				mg/L	90	Standard
	Rb	85	626.7	7.4				ug/L	23	Standard
	Y	89	465817.4	0.7				ug/L	448510	Standard
>	**Rh**	103	43.3	24.0				ug/L	15	Standard
	Mo	98	284447.8	1.6	**100.4857**	3.626	3.6	ug/L	11	Standard
	Ag	107	346501.9	1.0	**49.2778**	1.406	2.9	ug/L	134	Standard
	Cd	111	109630.9	1.1	**48.9766**	1.535	3.1	mg/L	7	Standard
	Cd	114	266600.1	1.2	**48.9905**	1.146	2.3	ug/L	16	Standard
>	**In**	115	640190.8	2.0				ug/L	611117	Standard
	Sn	118	332613.3	1.3	**53.6148**	1.570	2.9	ug/L	518	Standard
	Sb	123	238639.0	2.2	**48.2943**	2.007	4.2	ug/L	127	Standard
	Ba	135	104330.0	1.0	**48.2232**	0.702	1.5	ug/L	45	Standard
	Ce	140	120.0	15.0				ug/L	298	Standard
>	**Tb**	159	1038346.1	1.3				ug/L	977942	Standard
	Ho	165	16.7	34.6				ug/L	12	Standard
	Tl	203	439433.3	1.3	**49.1017**	0.655	1.3	ug/L	31	Standard
	Tl	205	379171.3	1.2	**48.7970**	1.107	2.3	ug/L	20	Standard
	Pb	206	282861.9	1.4	**49.2211**	1.133	2.3	ug/L	286	Standard
	Pb	207	247275.0	1.2	**47.7052**	1.056	2.2	ug/L	257	Standard
	Pb	208	979823.4	0.2	**48.8569**	0.617	1.3	ug/L	956	Standard
	U	238	361651.7	0.7	**47.5263**	0.839	1.8	ug/L	17	Standard
>	**Bi**	209	591677.6	1.0				ug/L	557080	Standard

Sample ID: QC Std 1
Report Date/Time: Tuesday, June 28, 2016 13:21:14
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.0435**	17.387	173.1	mg/L	0	Standard
Mg	24	4277.3	3.5	**5.2835**	0.147	2.8	mg/L	32	Standard
K	39	1651.8	3.6	**5.0931**	0.136	2.7	mg/L	30	Standard
Ca	43	81.7	30.2	**6.2885**	3.508	55.8	mg/L	48	Standard
Fe	54	5996.6	1.6	**4.9548**	0.129	2.6	mg/L	159	Standard
Fe	57	1723.4	6.4	**4.8356**	0.391	8.1	mg/L	195	Standard
Sc-1	45	42151.9	1.0				mg/L	39543	Standard
Cl	35	29861.4	1.9				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1713.4	14.2				ug/L	1647	Standard
P	31	18880.1	1.3				ug/L	15231	Standard
S	34	2316.8	4.8				ug/L	1845	Standard
Sr	88	70.0	35.7				ug/L	70	Standard
C	12	156.7	32.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	46.2	33.1				mg/L	12	Standard
Ho-1	165	16.7	34.6				mg/L	12	Standard
Er	166	10.0					mg/L	20	Standard
I	127	1246.7	11.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	100.857		
Al	27	102.543		
Sc	45			
Ti	47	103.222		
V	51	99.495		
Cr	52	99.958		
Cr	53			
Mn	55	100.358		
Co	59	100.312		
Ni	60	99.375		
Cu	65	100.213		
Zn	66	101.397		
Ge	72		104.263	
As	75	100.627		
Se	82	99.807		
Se-1	77			
Ga	71			

Sample ID: QC Std 1
Report Date/Time: Tuesday, June 28, 2016 13:21:14
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	100.486		
Ag	107	98.556		
Cd	111	97.953		
Cd	114			
In	115		104.758	
Sn	118	107.230		
Sb	123	96.589		
Ba	135	96.446		
Ce	140			
Tb	159			
Ho	165			
Tl	203	98.203		
Tl	205			
Pb	206	98.442		
Pb	207	95.410		
Pb	208	97.714		
U	238	95.053		
Bi	209		106.210	
Na	23	200.869		
Mg	24	105.670		
K	39	101.861		
Ca	43	125.771		
Fe	54	99.097		
Fe	57	96.711		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 1	Na	23	
QC Std 1	Ca	43	

Sample ID: QC Std 1
Report Date/Time: Tuesday, June 28, 2016 13:21:14
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 2
Sample Date/Time: Tuesday, June 28, 2016 13:22:10
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81098.1	0.9				ug/L	81760	Standard
	Be	9	35.0	14.3	**0.0258**	0.005	18.6	ug/L	17	Standard
	Al	27	2568.6	16.6	**0.0024**	0.003	117.4	ug/L	1410	Standard
	Sc	45	40706.3	0.9				ug/L	39543	Standard
	Ti	47	40.7	7.5	**-0.0093**	0.010	110.5	ug/L	39	Standard
	V	51	1184.2	4.7	**-0.0443**	0.005	11.8	ug/L	1415	Standard
	Cr	52	6261.3	1.2	**-0.1999**	0.018	9.1	ug/L	6946	Standard
	Cr	53	338.3	3.7	**0.0258**	0.014	55.9	ug/L	328	Standard
	Mn	55	842.4	4.2	**-0.0247**	0.005	18.3	ug/L	823	Standard
	Co	59	286.0	9.3	**0.0074**	0.003	40.3	ug/L	252	Standard
	Ni	60	199.7	6.5	**-0.0040**	0.005	124.9	ug/L	192	Standard
	Cu	65	142.3	4.7	**-0.0195**	0.003	13.5	ug/L	141	Standard
	Zn	66	529.7	2.7	**-0.3672**	0.009	2.6	ug/L	491	Standard
>	**Ge**	72	540586.6	1.6				ug/L	515087	Standard
	As	75	16.4	99.5	**0.0543**	0.014	25.8	ug/L	20	Standard
	Se	82	18.7	7.6	**0.0353**	0.014	40.6	ug/L	15	Standard
	Se-1	77	50.3	10.9	**-0.0819**	0.075	91.1	ug/L	48	Standard
>	**Ga**	71	86.7	18.5				mg/L	90	Standard
	Rb	85	13.3	78.1				ug/L	23	Standard
	Y	89	464568.8	2.3				ug/L	448510	Standard
>	**Rh**	103	18.3	15.7				ug/L	15	Standard
	Mo	98	439.4	21.7	**0.1377**	0.033	23.9	ug/L	11	Standard
	Ag	107	149.7	25.4	**0.0053**	0.005	102.3	ug/L	134	Standard
	Cd	111	15.2	39.6	**0.0040**	0.003	67.9	mg/L	7	Standard
	Cd	114	45.8	32.1	**0.0036**	0.003	74.9	ug/L	16	Standard
>	**In**	115	633054.7	0.7				ug/L	611117	Standard
	Sn	118	1025.0	34.7	**0.0655**	0.058	89.1	ug/L	518	Standard
	Sb	123	2042.3	25.6	**0.4053**	0.108	26.7	ug/L	127	Standard
	Ba	135	49.7	4.2	**-0.0019**	0.001	60.7	ug/L	45	Standard
	Ce	140	36.7	20.8				ug/L	298	Standard
>	**Tb**	159	1001668.5	0.7				ug/L	977942	Standard
	Ho	165	10.0					ug/L	12	Standard
	Tl	203	78.0	41.6	**0.0041**	0.004	88.8	ug/L	31	Standard
	Tl	205	58.3	47.2	**0.0054**	0.004	66.7	ug/L	20	Standard
	Pb	206	315.3	11.1	**-0.0075**	0.006	81.9	ug/L	286	Standard
	Pb	207	291.3	4.8	**-0.0005**	0.003	477.4	ug/L	257	Standard
	Pb	208	1079.0	5.9	**-0.0054**	0.003	55.4	ug/L	956	Standard
	U	238	63.3	42.9	**0.0055**	0.004	65.2	ug/L	17	Standard
>	**Bi**	209	578819.3	0.5				ug/L	557080	Standard

Sample ID: QC Std 2
Report Date/Time: Tuesday, June 28, 2016 13:24:27
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	25.0	40.0	**-0.0249**	0.013	50.9	mg/L	32	Standard
K	39	31.7	9.1	**-0.0375**	0.009	24.5	mg/L	30	Standard
Ca	43	41.7	36.7	**0.7901**	2.216	280.5	mg/L	48	Standard
Fe	54	146.0	8.6	**-0.0087**	0.010	113.7	mg/L	159	Standard
Fe	57	195.0	25.3	**0.0139**	0.156	1122.4	mg/L	195	Standard
Sc-1	45	40706.3	0.9				mg/L	39543	Standard
Cl	35	29554.8	2.8				ug/L	27642	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	1693.4	6.8				ug/L	1647	Standard
P	31	17536.8	2.2				ug/L	15231	Standard
S	34	2233.5	3.0				ug/L	1845	Standard
Sr	88	66.7	15.6				ug/L	70	Standard
C	12	96.7	33.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	12.7	49.8				mg/L	12	Standard
Ho-1	165	10.0					mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	1260.1	15.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		104.951	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 2
Report Date/Time: Tuesday, June 28, 2016 13:24:27
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	103.590
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	103.902
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 2	Sb	123	
QC Std 2	Ca	43	

Sample ID: QC Std 2
Report Date/Time: Tuesday, June 28, 2016 13:24:27
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 3
Sample Date/Time: Tuesday, June 28, 2016 13:25:23
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79486.3	1.5				ug/L	81760	Standard
	Be	9	236.7	23.6	**0.2350**	0.054	23.1	ug/L	17	Standard
	Al	27	1635.1	14.9	**-0.0035**	0.001	42.0	ug/L	1410	Standard
	Sc	45	39860.7	2.7				ug/L	39543	Standard
	Ti	47	39.7	2.9	**-0.0059**	0.005	91.5	ug/L	39	Standard
	V	51	3750.6	3.6	**0.3321**	0.019	5.7	ug/L	1415	Standard
	Cr	52	11750.6	2.2	**0.6283**	0.044	7.0	ug/L	6946	Standard
	Cr	53	1020.0	3.7	**0.8185**	0.038	4.7	ug/L	328	Standard
	Mn	55	3800.5	0.2	**0.4368**	0.004	0.8	ug/L	823	Standard
	Co	59	2925.3	4.6	**0.3843**	0.020	5.3	ug/L	252	Standard
	Ni	60	3080.3	1.6	**1.5280**	0.033	2.2	ug/L	192	Standard
	Cu	65	1617.4	2.7	**0.7591**	0.027	3.5	ug/L	141	Standard
	Zn	66	7652.9	2.4	**6.1434**	0.191	3.1	ug/L	491	Standard
>	**Ge**	72	515947.7	0.4				ug/L	515087	Standard
	As	75	426.1	10.9	**0.4186**	0.040	9.6	ug/L	20	Standard
	Se	82	65.5	20.2	**0.4487**	0.114	25.5	ug/L	15	Standard
	Se-1	77	77.0	13.5	**0.3017**	0.140	46.4	ug/L	48	Standard
>	**Ga**	71	75.0	29.1				mg/L	90	Standard
	Rb	85	11.7	49.5				ug/L	23	Standard
	Y	89	454544.4	2.9				ug/L	448510	Standard
>	**Rh**	103	11.7	65.5				ug/L	15	Standard
	Mo	98	144.3	6.8	**0.0328**	0.004	11.5	ug/L	11	Standard
	Ag	107	2677.9	4.8	**0.3715**	0.022	5.8	ug/L	134	Standard
	Cd	111	522.1	8.6	**0.2344**	0.020	8.5	mg/L	7	Standard
	Cd	114	1313.6	1.7	**0.2408**	0.006	2.4	ug/L	16	Standard
>	**In**	115	628989.1	0.7				ug/L	611117	Standard
	Sn	118	611.7	1.9	**-0.0014**	0.001	96.1	ug/L	518	Standard
	Sb	123	2074.6	8.0	**0.4143**	0.033	7.9	ug/L	127	Standard
	Ba	135	1500.7	4.3	**0.6813**	0.036	5.2	ug/L	45	Standard
	Ce	140	23.3	24.7				ug/L	298	Standard
>	**Tb**	159	1002704.6	1.0				ug/L	977942	Standard
	Ho	165	21.7	35.3				ug/L	12	Standard
	Tl	203	715.4	2.1	**0.0768**	0.002	2.1	ug/L	31	Standard
	Tl	205	653.3	9.4	**0.0835**	0.008	10.1	ug/L	20	Standard
	Pb	206	1358.4	2.4	**0.1778**	0.007	3.8	ug/L	286	Standard
	Pb	207	1172.0	3.3	**0.1729**	0.010	5.8	ug/L	257	Standard
	Pb	208	4784.3	0.9	**0.1832**	0.004	2.4	ug/L	956	Standard
	U	238	2840.3	2.6	**0.3778**	0.014	3.6	ug/L	17	Standard
>	**Bi**	209	580093.5	1.0				ug/L	557080	Standard

Sample ID: QC Std 3
Report Date/Time: Tuesday, June 28, 2016 13:27:40
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	33.3	22.9	**-0.0129**	0.011	85.9	mg/L	32	Standard
K	39	38.3	45.8	**-0.0130**	0.059	454.3	mg/L	30	Standard
Ca	43	38.3	41.9	**0.4748**	2.540	535.0	mg/L	48	Standard
Fe	54	103.9	16.3	**-0.0436**	0.016	37.7	mg/L	159	Standard
Fe	57	210.0	19.5	**0.0806**	0.150	185.6	mg/L	195	Standard
> **Sc-1**	45	39860.7	2.7				mg/L	39543	Standard
Cl	35	28271.0	1.2				ug/L	27642	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	1630.1	3.7				ug/L	1647	Standard
P	31	12470.1	4.9				ug/L	15231	Standard
S	34	2090.1	6.4				ug/L	1845	Standard
Sr	88	61.7	72.7				ug/L	70	Standard
C	12	146.7	7.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	26.0	116.4				mg/L	12	Standard
Ho-1	165	21.7	35.3				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	1031.7	25.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	117.475		
Al	27	-0.351		
Sc	45			
Ti	47			
V	51	83.021		
Cr	52	78.533		
Cr	53			
Mn	55	87.355		
Co	59	96.082		
Ni	60	95.499		
Cu	65	94.881		
Zn	66	98.294		
> Ge	72		100.167	
As	75	104.644		
Se	82	112.180		
Se-1	77			
> Ga	71			

Sample ID: QC Std 3
Report Date/Time: Tuesday, June 28, 2016 13:27:40
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	92.871		
Cd	111	97.662		
Cd	114			
> In	115			102.925
Sn	118			
Sb	123	103.570		
Ba	135	90.845		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	95.941		
Tl	205			
Pb	206			
Pb	207			
Pb	208	91.584		
U	238	94.451		
> Bi	209			104.131
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 3	Al	27	

Sample ID: QC Std 3
Report Date/Time: Tuesday, June 28, 2016 13:27:40
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Tuesday, June 28, 2016 13:28:34
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85379.9	0.7				ug/L	81760	Standard
	Be	9	28.3	44.4	**0.0176**	0.012	68.3	ug/L	17	Standard
	Al	27	7582301.1	0.5	**47.0918**	0.449	1.0	ug/L	1410	Standard
	Sc	45	42481.1	0.8				ug/L	39543	Standard
	Ti	47	26924.2	1.2	**102.1971**	2.644	2.6	ug/L	39	Standard
	V	51	1177.2	11.7	**-0.0466**	0.017	37.5	ug/L	1415	Standard
	Cr	52	7440.5	1.6	**-0.0468**	0.011	24.4	ug/L	6946	Standard
	Cr	53	2938.6	8.1	**2.8263**	0.310	11.0	ug/L	328	Standard
	Mn	55	1484.1	0.5	**0.0680**	0.007	9.9	ug/L	823	Standard
	Co	59	438.3	7.6	**0.0276**	0.004	13.4	ug/L	252	Standard
	Ni	60	870.4	2.5	**0.3318**	0.004	1.2	ug/L	192	Standard
	Cu	65	647.0	5.3	**0.2309**	0.010	4.5	ug/L	141	Standard
	Zn	66	2006.5	1.8	**0.9027**	0.019	2.1	ug/L	491	Standard
>	**Ge**	72	545025.5	2.5				ug/L	515087	Standard
	As	75	-62.2	34.4	**-0.0118**	0.017	144.3	ug/L	20	Standard
	Se	82	23.6	6.2	**0.0746**	0.009	12.7	ug/L	15	Standard
	Se-1	77	231.7	5.9	**2.1902**	0.222	10.1	ug/L	48	Standard
>	**Ga**	71	171.7	17.1				mg/L	90	Standard
	Rb	85	1090.0	4.4				ug/L	23	Standard
	Y	89	452797.1	1.4				ug/L	448510	Standard
>	**Rh**	103	10.0	100.0				ug/L	15	Standard
	Mo	98	263118.9	1.1	**94.9055**	0.578	0.6	ug/L	11	Standard
	Ag	107	286.7	3.7	**0.0254**	0.002	6.6	ug/L	134	Standard
	Cd	111	-226.3	7.3	**-0.1062**	0.009	8.1	mg/L	7	Standard
	Cd	114	466.4	13.0	**0.0826**	0.011	12.8	ug/L	16	Standard
>	**In**	115	626700.1	1.0				ug/L	611117	Standard
	Sn	118	941.7	5.7	**0.0533**	0.007	13.8	ug/L	518	Standard
	Sb	123	466.3	29.2	**0.0836**	0.028	33.8	ug/L	127	Standard
	Ba	135	616.7	6.0	**0.2662**	0.017	6.2	ug/L	45	Standard
	Ce	140	233.3	17.2				ug/L	298	Standard
>	**Tb**	159	1019781.5	0.4				ug/L	977942	Standard
	Ho	165	28.3	27.0				ug/L	12	Standard
	Tl	203	270.7	14.6	**0.0257**	0.004	17.2	ug/L	31	Standard
	Tl	205	205.0	17.1	**0.0243**	0.005	18.8	ug/L	20	Standard
	Pb	206	568.7	4.5	**0.0362**	0.004	12.0	ug/L	286	Standard
	Pb	207	508.0	2.4	**0.0408**	0.002	5.7	ug/L	257	Standard
	Pb	208	1902.0	8.6	**0.0352**	0.008	23.2	ug/L	956	Standard
	U	238	71.7	54.2	**0.0065**	0.005	78.6	ug/L	17	Standard
>	**Bi**	209	587460.1	0.4				ug/L	557080	Standard

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 13:30:51
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**29.8288**	29.665	99.5	mg/L	0	Standard
Mg	24	10050.0	1.2	**12.3961**	0.163	1.3	mg/L	32	Standard
K	39	1535.1	9.7	**4.6887**	0.503	10.7	mg/L	30	Standard
Ca	43	211.7	14.4	**24.5987**	4.529	18.4	mg/L	48	Standard
Fe	54	14184.2	1.5	**11.8132**	0.271	2.3	mg/L	159	Standard
Fe	57	3883.8	1.7	**11.5795**	0.109	0.9	mg/L	195	Standard
Sc-1	45	42481.1	0.8				mg/L	39543	Standard
Cl	35	27442.8	1.7				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1816.8	6.7				ug/L	1647	Standard
P	31	16744.2	0.2				ug/L	15231	Standard
S	34	1956.8	7.4				ug/L	1845	Standard
Sr	88	48.3	53.1				ug/L	70	Standard
C	12	220.0	25.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	56.7	27.0				mg/L	0	Standard
Dy	164	24.9	47.6				mg/L	12	Standard
Ho-1	165	28.3	27.0				mg/L	12	Standard
Er	166	36.7	56.8				mg/L	20	Standard
I	127	1896.8	6.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27	0.942		
Sc	45			
Ti	47	102.197		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		105.812	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 13:30:51
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	94.905		
Ag	107			
Cd	111			
Cd	114			
In	115			102.550
Sn	118			
Sb	123			
Ba	135			
Ce	140			
Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
Bi	209			105.453
Na	23	238.631		
Mg	24	247.921		
K	39	93.774		
Ca	43	163.991		
Fe	54	94.506		
Fe	57	92.636		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Na	23	
QC Std 4	Mg	24	

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 13:30:51
Page 3

Approved: June 29, 2016

QC Std 4	Ca	43

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 13:30:51
Page 4

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Tuesday, June 28, 2016 13:31:45
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87031.0	2.2				ug/L	81760	Standard
	Be	9	104600.0	3.1	**98.9445**	2.636	2.7	ug/L	17	Standard
	Al	27	8235764.9	2.8	**50.1881**	1.494	3.0	ug/L	1410	Standard
	Sc	45	40814.9	1.3				ug/L	39543	Standard
	Ti	47	28557.9	2.6	**110.3128**	3.806	3.5	ug/L	39	Standard
	V	51	698774.9	0.6	**96.5097**	2.203	2.3	ug/L	1415	Standard
	Cr	52	706026.0	2.3	**96.5109**	2.799	2.9	ug/L	6946	Standard
	Cr	53	91691.9	3.7	**100.1429**	4.249	4.2	ug/L	328	Standard
	Mn	55	659133.1	0.8	**97.6603**	2.181	2.2	ug/L	823	Standard
	Co	59	714164.5	1.5	**97.7493**	2.387	2.4	ug/L	252	Standard
	Ni	60	188241.9	2.2	**96.0473**	2.835	3.0	ug/L	192	Standard
	Cu	65	190773.3	1.3	**96.4911**	2.052	2.1	ug/L	141	Standard
	Zn	66	115799.2	0.5	**100.8065**	2.067	2.1	ug/L	491	Standard
>	**Ge**	72	535562.3	1.7				ug/L	515087	Standard
	As	75	115513.0	0.8	**98.8727**	1.879	1.9	ug/L	20	Standard
	Se	82	11807.8	0.8	**98.5471**	2.365	2.4	ug/L	15	Standard
	Se-1	77	7973.8	0.6	**101.0979**	1.926	1.9	ug/L	48	Standard
>	**Ga**	71	195.0	2.6				mg/L	90	Standard
	Rb	85	900.0	5.1				ug/L	23	Standard
	Y	89	467512.1	1.1				ug/L	448510	Standard
>	**Rh**	103	41.7	61.6				ug/L	15	Standard
	Mo	98	278616.6	1.0	**98.3192**	1.671	1.7	ug/L	11	Standard
	Ag	107	629653.8	2.5	**89.4581**	1.609	1.8	ug/L	134	Standard
	Cd	111	214179.9	0.7	**95.5849**	1.082	1.1	mg/L	7	Standard
	Cd	114	513207.5	0.9	**94.2229**	0.578	0.6	ug/L	16	Standard
>	**In**	115	640649.3	1.3				ug/L	611117	Standard
	Sn	118	1085.0	2.4	**0.0732**	0.006	8.8	ug/L	518	Standard
	Sb	123	468645.5	2.2	**94.7292**	1.132	1.2	ug/L	127	Standard
	Ba	135	204538.3	1.7	**94.4814**	1.088	1.2	ug/L	45	Standard
	Ce	140	98.3	7.8				ug/L	298	Standard
>	**Tb**	159	1030592.0	2.2				ug/L	977942	Standard
	Ho	165	16.7	45.8				ug/L	12	Standard
	Tl	203	850272.7	1.9	**96.3847**	2.377	2.5	ug/L	31	Standard
	Tl	205	753805.3	1.5	**98.4067**	2.429	2.5	ug/L	20	Standard
	Pb	206	560444.8	0.3	**98.9799**	0.961	1.0	ug/L	286	Standard
	Pb	207	485626.8	0.4	**95.0861**	1.259	1.3	ug/L	257	Standard
	Pb	208	1940003.2	1.2	**98.1909**	2.155	2.2	ug/L	956	Standard
	U	238	731621.6	1.7	**97.5330**	2.488	2.6	ug/L	17	Standard
>	**Bi**	209	583301.0	1.0				ug/L	557080	Standard

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 13:34:02
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**30.9951**	30.718	99.1	mg/L	0	Standard
Mg	24	10642.1	4.7	**13.6704**	0.697	5.1	mg/L	32	Standard
K	39	1555.1	3.7	**4.9498**	0.207	4.2	mg/L	30	Standard
Ca	43	191.7	19.9	**22.8900**	5.801	25.3	mg/L	48	Standard
Fe	54	15478.2	1.2	**13.4350**	0.150	1.1	mg/L	159	Standard
Fe	57	4137.2	0.8	**12.9082**	0.151	1.2	mg/L	195	Standard
> **Sc-1**	45	40814.9	1.3				mg/L	39543	Standard
Cl	35	27689.3	2.5				ug/L	27642	Standard
Kr	83	2.0	100.0				ug/L	2	Standard
Br	81	1876.8	10.9				ug/L	1647	Standard
P	31	15799.9	3.3				ug/L	15231	Standard
S	34	1943.5	5.2				ug/L	1845	Standard
Sr	88	65.0	27.7				ug/L	70	Standard
C	12	256.7	13.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	15.9	127.8				mg/L	12	Standard
Ho-1	165	16.7	45.8				mg/L	12	Standard
Er	166	16.7	69.3				mg/L	20	Standard
I	127	1231.7	9.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	98.944		
Al	27	1.004		
Sc	45			
Ti	47	110.313		
V	51	96.510		
Cr	52	96.511		
Cr	53			
Mn	55	97.660		
Co	59	97.749		
Ni	60	96.047		
Cu	65	96.491		
Zn	66	100.807		
> Ge	72		103.975	
As	75	98.873		
Se	82	98.547		
Se-1	77			
> Ga	71			

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 13:34:02
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	98.319		
Ag	107	89.458		
Cd	111	95.585		
Cd	114			
In	115		104.833	
Sn	118			
Sb	123	94.729		
Ba	135	94.481		
Ce	140			
Tb	159			
Ho	165			
Tl	203	96.385		
Tl	205			
Pb	206			
Pb	207			
Pb	208	98.191		
U	238	97.533		
Bi	209		104.707	
Na	23	247.961		
Mg	24	273.408		
K	39	98.997		
Ca	43	152.600		
Fe	54	107.480		
Fe	57	103.265		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Na	23	
QC Std 5	Mg	24	

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 13:34:02
Page 3

Approved: June 29, 2016

QC Std 5	Ca	43

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 13:34:02
Page 4

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 13:34:59
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85572.8	3.1				ug/L	81760	Standard
	Be	9	51348.6	3.0	**49.4131**	1.801	3.6	ug/L	17	Standard
	Al	27	7869928.8	0.7	**48.7904**	1.162	2.4	ug/L	1410	Standard
	Sc	45	41500.1	1.3				ug/L	39543	Standard
	Ti	47	25257.0	1.4	**98.1904**	2.232	2.3	ug/L	39	Standard
	V	51	359717.6	0.9	**49.9155**	1.252	2.5	ug/L	1415	Standard
	Cr	52	362752.3	1.0	**49.4079**	0.954	1.9	ug/L	6946	Standard
	Cr	53	45483.2	1.3	**49.8439**	1.589	3.2	ug/L	328	Standard
	Mn	55	333977.4	0.1	**49.7416**	1.012	2.0	ug/L	823	Standard
	Co	59	371364.3	0.2	**51.1559**	1.087	2.1	ug/L	252	Standard
	Ni	60	96391.0	0.8	**49.4567**	0.760	1.5	ug/L	192	Standard
	Cu	65	97494.7	0.8	**49.5943**	0.620	1.2	ug/L	141	Standard
	Zn	66	56948.5	0.8	**49.4853**	0.762	1.5	ug/L	491	Standard
>	**Ge**	72	532023.1	2.0				ug/L	515087	Standard
	As	75	57437.4	0.6	**49.5161**	1.180	2.4	ug/L	20	Standard
	Se	82	5946.0	0.9	**49.9013**	1.415	2.8	ug/L	15	Standard
	Se-1	77	4004.5	1.6	**50.7471**	0.199	0.4	ug/L	48	Standard
>	**Ga**	71	93.3	16.4				mg/L	90	Standard
	Rb	85	881.7	8.3				ug/L	23	Standard
	Y	89	456900.1	0.4				ug/L	448510	Standard
>	**Rh**	103	26.7	60.3				ug/L	15	Standard
	Mo	98	274645.7	0.6	**96.2670**	1.141	1.2	ug/L	11	Standard
	Ag	107	344865.6	0.2	**48.6675**	0.302	0.6	ug/L	134	Standard
	Cd	111	109918.0	0.6	**48.7237**	0.167	0.3	mg/L	7	Standard
	Cd	114	266820.9	0.9	**48.6598**	0.652	1.3	ug/L	16	Standard
>	**In**	115	644937.8	0.8				ug/L	611117	Standard
	Sn	118	304766.5	1.2	**48.7367**	0.172	0.4	ug/L	518	Standard
	Sb	123	245072.6	0.1	**49.2072**	0.383	0.8	ug/L	127	Standard
	Ba	135	106223.8	1.1	**48.7299**	0.483	1.0	ug/L	45	Standard
	Ce	140	76.7	21.0				ug/L	298	Standard
>	**Tb**	159	1040424.8	0.9				ug/L	977942	Standard
	Ho	165	11.7	89.2				ug/L	12	Standard
	Tl	203	436254.3	0.9	**48.8621**	1.048	2.1	ug/L	31	Standard
	Tl	205	382952.9	1.1	**49.3898**	0.698	1.4	ug/L	20	Standard
	Pb	206	278752.3	0.9	**48.6168**	1.163	2.4	ug/L	286	Standard
	Pb	207	252906.4	1.0	**48.9055**	1.215	2.5	ug/L	257	Standard
	Pb	208	993060.2	1.0	**49.6339**	1.159	2.3	ug/L	956	Standard
	U	238	369537.8	0.4	**48.6731**	0.911	1.9	ug/L	17	Standard
>	**Bi**	209	590375.8	1.5				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 13:37:15
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4017.2	5.2	**5.0405**	0.318	6.3	mg/L	32	Standard
K	39	1600.1	5.0	**5.0128**	0.329	6.6	mg/L	30	Standard
Ca	43	98.3	19.3	**8.9105**	2.944	33.0	mg/L	48	Standard
Fe	54	5965.0	1.0	**5.0068**	0.016	0.3	mg/L	159	Standard
Fe	57	1785.1	5.1	**5.1208**	0.370	7.2	mg/L	195	Standard
> **Sc-1**	45	41500.1	1.3				mg/L	39543	Standard
Cl	35	29160.0	1.9				ug/L	27642	Standard
Kr	83	3.3	45.8				ug/L	2	Standard
Br	81	1670.1	8.2				ug/L	1647	Standard
P	31	17116.3	2.7				ug/L	15231	Standard
S	34	2186.8	7.3				ug/L	1845	Standard
Sr	88	41.7	34.6				ug/L	70	Standard
C	12	130.0	42.8				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	23.3	24.7				mg/L	12	Standard
Ho-1	165	11.7	89.2				mg/L	12	Standard
Er	166	0.0					mg/L	20	Standard
I	127	1351.7	14.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	98.826		
Al	27	97.581		
Sc	45			
Ti	47	98.190		
V	51	99.831		
Cr	52	98.816		
Cr	53			
Mn	55	99.483		
Co	59	102.312		
Ni	60	98.913		
Cu	65	99.189		
Zn	66	98.971		
> Ge	72		103.288	
As	75	99.032		
Se	82	99.803		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 13:37:15
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	96.267		
Ag	107	97.335		
Cd	111	97.447		
Cd	114			
In	115			105.534
Sn	118	97.473		
Sb	123	98.414		
Ba	135	97.460		
Ce	140			
Tb	159			
Ho	165			
Tl	203	97.724		
Tl	205			
Pb	206			
Pb	207			
Pb	208	99.268		
U	238	97.346		
Bi	209			105.977
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 13:37:15
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 13:38:10
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82763.8	0.7				ug/L	81760	Standard
	Be	9	40.0	57.3	**0.0300**	0.023	75.9	ug/L	17	Standard
	Al	27	4257.4	63.2	**0.0129**	0.017	134.4	ug/L	1410	Standard
	Sc	45	40355.3	1.7				ug/L	39543	Standard
	Ti	47	44.3	23.5	**0.0085**	0.041	476.8	ug/L	39	Standard
	V	51	1283.7	14.1	**-0.0268**	0.025	92.2	ug/L	1415	Standard
	Cr	52	6366.3	1.6	**-0.1663**	0.013	7.9	ug/L	6946	Standard
	Cr	53	461.7	3.5	**0.1707**	0.014	8.3	ug/L	328	Standard
	Mn	55	939.4	15.6	**-0.0074**	0.022	294.6	ug/L	823	Standard
	Co	59	428.3	39.2	**0.0280**	0.023	82.9	ug/L	252	Standard
	Ni	60	218.7	25.5	**0.0081**	0.029	354.3	ug/L	192	Standard
	Cu	65	184.3	18.2	**0.0036**	0.017	469.2	ug/L	141	Standard
	Zn	66	515.3	2.3	**-0.3696**	0.014	3.7	ug/L	491	Standard
>	**Ge**	72	528838.2	0.8				ug/L	515087	Standard
	As	75	8.3	324.4	**0.0475**	0.023	49.1	ug/L	20	Standard
	Se	82	19.9	12.2	**0.0491**	0.021	43.7	ug/L	15	Standard
	Se-1	77	55.0	4.8	**-0.0079**	0.029	369.8	ug/L	48	Standard
>	**Ga**	71	56.7	39.8				mg/L	90	Standard
	Rb	85	10.0					ug/L	23	Standard
	Y	89	469796.8	1.6				ug/L	448510	Standard
>	**Rh**	103	20.0	25.0				ug/L	15	Standard
	Mo	98	452.3	25.3	**0.1439**	0.042	29.5	ug/L	11	Standard
	Ag	107	307.7	67.7	**0.0286**	0.031	108.2	ug/L	134	Standard
	Cd	111	45.2	95.8	**0.0178**	0.020	112.8	mg/L	7	Standard
	Cd	114	114.8	113.1	**0.0168**	0.025	147.3	ug/L	16	Standard
>	**In**	115	628024.0	1.4				ug/L	611117	Standard
	Sn	118	940.0	23.7	**0.0531**	0.039	73.1	ug/L	518	Standard
	Sb	123	1312.1	39.0	**0.2588**	0.110	42.3	ug/L	127	Standard
	Ba	135	79.3	75.1	**0.0125**	0.029	229.7	ug/L	45	Standard
	Ce	140	28.3	36.7				ug/L	298	Standard
>	**Tb**	159	1003248.8	1.6				ug/L	977942	Standard
	Ho	165	15.0	88.2				ug/L	12	Standard
	Tl	203	265.0	107.1	**0.0261**	0.033	127.7	ug/L	31	Standard
	Tl	205	163.3	84.5	**0.0196**	0.019	95.5	ug/L	20	Standard
	Pb	206	421.0	42.3	**0.0127**	0.033	261.4	ug/L	286	Standard
	Pb	207	362.0	44.3	**0.0147**	0.033	224.6	ug/L	257	Standard
	Pb	208	1358.7	30.8	**0.0100**	0.022	223.6	ug/L	956	Standard
	U	238	139.0	84.9	**0.0161**	0.016	101.8	ug/L	17	Standard
>	**Bi**	209	570467.7	1.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 13:40:27
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.3221**	17.870	173.1	mg/L	0	Standard
Mg	24	30.0	28.9	**-0.0179**	0.012	67.0	mg/L	32	Standard
K	39	31.7	24.1	**-0.0367**	0.024	65.4	mg/L	30	Standard
Ca	43	38.3	15.1	**0.3686**	0.952	258.2	mg/L	48	Standard
Fe	54	115.7	16.2	**-0.0344**	0.017	49.5	mg/L	159	Standard
Fe	57	198.3	8.1	**0.0320**	0.061	191.3	mg/L	195	Standard
> **Sc-1**	45	40355.3	1.7				mg/L	39543	Standard
Cl	35	29509.4	2.2				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	1783.4	4.3				ug/L	1647	Standard
P	31	17421.7	1.2				ug/L	15231	Standard
S	34	2170.2	3.8				ug/L	1845	Standard
Sr	88	55.0	15.7				ug/L	70	Standard
C	12	93.3	40.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	8.9	120.7				mg/L	12	Standard
Ho-1	165	15.0	88.2				mg/L	12	Standard
Er	166	23.3	65.5				mg/L	20	Standard
I	127	1233.4	14.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.670	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 13:40:27
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	102.767
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.403
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Sb	123	

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 13:40:27
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036203
Sample Date/Time: Tuesday, June 28, 2016 13:41:22
Number of Replicates: 3
Autosampler Position: 301
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83400.5	10.2				ug/L	81760	Standard
	Be	9	11.7	65.5	**0.0015**	0.006	425.0	ug/L	17	Standard
	Al	27	3017.0	39.5	**0.0053**	0.009	165.2	ug/L	1410	Standard
	Sc	45	40218.8	12.0				ug/L	39543	Standard
	Ti	47	37.0	23.6	**-0.0163**	0.022	135.1	ug/L	39	Standard
	V	51	613.8	4.6	**-0.1167**	0.012	10.1	ug/L	1415	Standard
	Cr	52	3481.1	12.6	**-0.5531**	0.019	3.4	ug/L	6946	Standard
	Cr	53	250.0	27.5	**-0.0570**	0.057	99.8	ug/L	328	Standard
	Mn	55	1041.4	26.2	**0.0117**	0.027	231.5	ug/L	823	Standard
	Co	59	254.7	5.9	**0.0055**	0.004	71.3	ug/L	252	Standard
	Ni	60	206.3	13.3	**0.0056**	0.009	153.1	ug/L	192	Standard
	Cu	65	131.7	19.5	**-0.0206**	0.013	65.4	ug/L	141	Standard
	Zn	66	531.7	22.4	**-0.3419**	0.069	20.1	ug/L	491	Standard
>	**Ge**	72	510441.4	10.5				ug/L	515087	Standard
	As	75	44.6	25.0	**0.0799**	0.006	7.8	ug/L	20	Standard
	Se	82	38.3	36.8	**0.2107**	0.093	44.1	ug/L	15	Standard
	Se-1	77	32.0	11.3	**-0.2872**	0.062	21.6	ug/L	48	Standard
>	**Ga**	71	61.7	4.7				mg/L	90	Standard
	Rb	85	33.3	37.7				ug/L	23	Standard
	Y	89	399549.3	20.8				ug/L	448510	Standard
>	**Rh**	103	11.7	65.5				ug/L	15	Standard
	Mo	98	155.9	22.2	**0.0482**	0.023	47.8	ug/L	11	Standard
	Ag	107	121.0	24.4	**0.0044**	0.005	118.2	ug/L	134	Standard
	Cd	111	13.4	45.5	**0.0046**	0.004	96.5	mg/L	7	Standard
	Cd	114	43.9	10.7	**0.0047**	0.001	14.6	ug/L	16	Standard
>	**In**	115	541793.5	15.9				ug/L	611117	Standard
	Sn	118	351.7	31.7	**-0.0361**	0.012	32.5	ug/L	518	Standard
	Sb	123	283.1	30.2	**0.0538**	0.012	21.7	ug/L	127	Standard
	Ba	135	37.0	32.8	**-0.0052**	0.004	82.4	ug/L	45	Standard
	Ce	140	30.0	28.9				ug/L	298	Standard
>	**Tb**	159	879683.0	16.9				ug/L	977942	Standard
	Ho	165	8.3	34.6				ug/L	12	Standard
	Tl	203	99.7	75.1	**0.0097**	0.013	137.9	ug/L	31	Standard
	Tl	205	85.0	86.7	**0.0121**	0.015	120.6	ug/L	20	Standard
	Pb	206	313.0	1.9	**0.0011**	0.012	1036.8	ug/L	286	Standard
	Pb	207	266.0	5.5	**0.0034**	0.015	442.4	ug/L	257	Standard
	Pb	208	1233.7	12.3	**0.0133**	0.020	150.4	ug/L	956	Standard
	U	238	39.3	36.7	**0.0031**	0.002	60.3	ug/L	17	Standard
>	**Bi**	209	508733.8	16.9				ug/L	557080	Standard

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:43:39
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	21.7	13.3	**-0.0282**	0.008	27.2	mg/L	32	Standard
K	39	25.0	34.6	**-0.0591**	0.022	36.5	mg/L	30	Standard
Ca	43	40.0	12.5	**0.7483**	1.524	203.6	mg/L	48	Standard
Fe	54	113.2	25.1	**-0.0334**	0.039	117.8	mg/L	159	Standard
Fe	57	238.3	14.7	**0.1717**	0.127	73.7	mg/L	195	Standard
> **Sc-1**	45	40218.8	12.0				mg/L	39543	Standard
Cl	35	13831.7	10.0				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	810.0	8.9				ug/L	1647	Standard
P	31	10527.1	24.9				ug/L	15231	Standard
S	34	970.0	6.2				ug/L	1845	Standard
Sr	88	66.7	11.5				ug/L	70	Standard
C	12	110.0	9.1				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	12.4	124.8				mg/L	12	Standard
Ho-1	165	8.3	34.6				mg/L	12	Standard
Er	166	20.0	50.0				mg/L	20	Standard
I	127	523.3	19.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.006	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.098	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:43:39
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		88.656
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		91.321
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Cr 52 Lower	Cr	52	

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:43:39
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036204
Sample Date/Time: Tuesday, June 28, 2016 13:44:34
Number of Replicates: 3
Autosampler Position: 302
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83225.5	7.5				ug/L	81760	Standard
	Be	9	31228.6	61.6	**31.9371**	21.971	68.8	ug/L	17	Standard
	Al	27	8471.1	48.6	**0.0411**	0.031	75.9	ug/L	1410	Standard
	Sc	45	40830.0	2.2				ug/L	39543	Standard
	Ti	47	34.3	11.8	**-0.0297**	0.018	59.1	ug/L	39	Standard
	V	51	224078.9	58.8	**31.3360**	18.302	58.4	ug/L	1415	Standard
	Cr	52	238015.8	57.7	**32.3999**	19.040	58.8	ug/L	6946	Standard
	Cr	53	29690.7	58.7	**32.7502**	19.168	58.5	ug/L	328	Standard
	Mn	55	213724.7	56.8	**32.1044**	18.055	56.2	ug/L	823	Standard
	Co	59	222487.2	58.7	**30.9454**	17.938	58.0	ug/L	252	Standard
	Ni	60	62090.3	57.3	**32.1492**	18.240	56.7	ug/L	192	Standard
	Cu	65	63757.0	57.0	**32.7355**	18.436	56.3	ug/L	141	Standard
	Zn	66	38227.9	57.9	**33.2873**	19.476	58.5	ug/L	491	Standard
>	**Ge**	72	525473.2	1.2				ug/L	515087	Standard
	As	75	36453.5	56.6	**31.7527**	17.695	55.7	ug/L	20	Standard
	Se	82	3780.3	55.0	**31.9968**	17.401	54.4	ug/L	15	Standard
	Se-1	77	2538.3	56.6	**32.2335**	18.366	57.0	ug/L	48	Standard
>	**Ga**	71	76.7	26.4				mg/L	90	Standard
	Rb	85	43.3	17.6				ug/L	23	Standard
	Y	89	474726.4	7.6				ug/L	448510	Standard
>	**Rh**	103	26.7	21.7				ug/L	15	Standard
	Mo	98	564.0	40.4	**0.1823**	0.089	49.0	ug/L	11	Standard
	Ag	107	212096.4	59.4	**30.6350**	19.274	62.9	ug/L	134	Standard
	Cd	111	68030.3	54.2	**30.8198**	17.865	58.0	mg/L	7	Standard
	Cd	114	165078.2	56.3	**30.7839**	18.487	60.1	ug/L	16	Standard
>	**In**	115	641067.2	4.2				ug/L	611117	Standard
	Sn	118	550.0	28.6	**-0.0125**	0.029	234.3	ug/L	518	Standard
	Sb	123	143570.7	60.9	**29.5172**	18.974	64.3	ug/L	127	Standard
	Ba	135	66133.4	54.2	**31.0035**	17.964	57.9	ug/L	45	Standard
	Ce	140	83.3	81.7				ug/L	298	Standard
>	**Tb**	159	1026938.7	4.8				ug/L	977942	Standard
	Ho	165	11.7	65.5				ug/L	12	Standard
	Tl	203	274073.1	56.4	**30.7538**	17.663	57.4	ug/L	31	Standard
	Tl	205	243234.0	54.4	**31.4411**	17.425	55.4	ug/L	20	Standard
	Pb	206	180330.9	55.5	**31.4753**	17.798	56.5	ug/L	286	Standard
	Pb	207	158009.7	55.1	**30.5868**	17.206	56.3	ug/L	257	Standard
	Pb	208	646417.5	54.5	**32.3514**	17.977	55.6	ug/L	956	Standard
	U	238	234720.5	55.2	**30.9733**	17.414	56.2	ug/L	17	Standard
>	**Bi**	209	588853.8	4.4				ug/L	557080	Standard

Sample ID: L1606036204
Report Date/Time: Tuesday, June 28, 2016 13:46:51
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.1717**	17.609	173.1	mg/L	0	Standard
Mg	24	26.7	10.8	**-0.0228**	0.004	16.0	mg/L	32	Standard
K	39	36.7	43.8	**-0.0214**	0.052	244.2	mg/L	30	Standard
Ca	43	30.0	66.7	**-0.9172**	2.942	320.7	mg/L	48	Standard
Fe	54	107.6	36.8	**-0.0427**	0.035	81.3	mg/L	159	Standard
Fe	57	220.0	9.9	**0.0952**	0.078	81.4	mg/L	195	Standard
> **Sc-1**	45	40830.0	2.2				mg/L	39543	Standard
Cl	35	20640.2	29.5				ug/L	27642	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	1230.1	37.8				ug/L	1647	Standard
P	31	11926.4	10.7				ug/L	15231	Standard
S	34	1536.7	28.4				ug/L	1845	Standard
Sr	88	66.7	11.5				ug/L	70	Standard
C	12	146.7	27.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	5.9	100.8				mg/L	12	Standard
Ho-1	165	11.7	65.5				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	785.0	24.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.792	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.016	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036204
Report Date/Time: Tuesday, June 28, 2016 13:46:51
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.901
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.704
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036203
Sample Date/Time: Tuesday, June 28, 2016 13:49:03
Number of Replicates: 3
Autosampler Position: 205
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90942.6	5.2				ug/L	81760	Standard
	Be	9	53.3	85.1	**0.0399**	0.045	111.9	ug/L	17	Standard
	Al	27	693137.8	2.1	**4.0335**	0.159	3.9	ug/L	1410	Standard
	Sc	45	39909.2	1.1				ug/L	39543	Standard
	Ti	47	362.3	4.5	**1.2955**	0.068	5.3	ug/L	39	Standard
	V	51	2278.9	9.8	**0.1220**	0.031	25.8	ug/L	1415	Standard
	Cr	52	6909.6	1.4	**-0.0622**	0.012	19.7	ug/L	6946	Standard
	Cr	53	1090.0	12.5	**0.9032**	0.155	17.1	ug/L	328	Standard
	Mn	55	5526773.3	1.7	**854.3657**	14.322	1.7	ug/L	823	Standard
	Co	59	9370.2	5.6	**1.3053**	0.073	5.6	ug/L	252	Standard
	Ni	60	2122.5	1.2	**1.0247**	0.015	1.4	ug/L	192	Standard
	Cu	65	2498.9	0.1	**1.2274**	0.002	0.2	ug/L	141	Standard
	Zn	66	6516.4	0.8	**5.1316**	0.039	0.8	ug/L	491	Standard
>	**Ge**	72	513880.5	0.2				ug/L	515087	Standard
	As	75	3228.7	3.4	**2.9189**	0.105	3.6	ug/L	20	Standard
	Se	82	922.2	2.2	**7.9103**	0.196	2.5	ug/L	15	Standard
	Se-1	77	155.0	10.7	**1.3431**	0.217	16.2	ug/L	48	Standard
>	**Ga**	71	388.3	17.5				mg/L	90	Standard
	Rb	85	68372.4	3.2				ug/L	23	Standard
	Y	89	447009.0	1.5				ug/L	448510	Standard
>	**Rh**	103	111.7	20.2				ug/L	15	Standard
	Mo	98	1200247.1	0.8	**451.3790**	8.919	2.0	ug/L	11	Standard
	Ag	107	153.3	18.9	**0.0070**	0.004	61.8	ug/L	134	Standard
	Cd	111	-1054.6	0.6	**-0.5044**	0.004	0.7	mg/L	7	Standard
	Cd	114	1800.6	2.5	**0.3474**	0.013	3.8	ug/L	16	Standard
>	**In**	115	601262.2	1.3				ug/L	611117	Standard
	Sn	118	788.4	4.6	**0.0336**	0.006	17.8	ug/L	518	Standard
	Sb	123	1882.7	34.8	**0.3916**	0.136	34.8	ug/L	127	Standard
	Ba	135	4520.7	1.0	**2.2009**	0.051	2.3	ug/L	45	Standard
	Ce	140	6458.0	2.1				ug/L	298	Standard
>	**Tb**	159	1000527.6	0.4				ug/L	977942	Standard
	Ho	165	130.0	21.4				ug/L	12	Standard
	Tl	203	285.7	2.5	**0.0299**	0.000	1.6	ug/L	31	Standard
	Tl	205	403.3	76.1	**0.0543**	0.043	79.9	ug/L	20	Standard
	Pb	206	899.7	2.3	**0.1068**	0.005	4.7	ug/L	286	Standard
	Pb	207	732.0	4.6	**0.0957**	0.009	9.0	ug/L	257	Standard
	Pb	208	3524.8	26.8	**0.1310**	0.052	40.0	ug/L	956	Standard
	U	238	352.3	76.1	**0.0475**	0.039	81.4	ug/L	17	Standard
>	**Bi**	209	544253.0	1.1				ug/L	557080	Standard

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:51:20
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.4757**	18.136	173.1	mg/L	0	Standard
Mg	24	122834.3	0.9	**161.9583**	0.885	0.5	mg/L	32	Standard
K	39	4865.8	5.3	**16.1536**	1.001	6.2	mg/L	30	Standard
Ca	43	186.7	17.4	**22.7843**	5.205	22.8	mg/L	48	Standard
Fe	54	927.7	8.4	**0.6952**	0.076	10.9	mg/L	159	Standard
Fe	57	461.7	8.7	**0.9208**	0.151	16.4	mg/L	195	Standard
> **Sc-1**	45	39909.2	1.1				mg/L	39543	Standard
Cl	35	47371.9	1.0				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	196320.5	1.2				ug/L	1647	Standard
P	31	16525.7	2.7				ug/L	15231	Standard
S	34	2001.8	3.8				ug/L	1845	Standard
Sr	88	358.3	4.3				ug/L	70	Standard
C	12	246.7	19.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	211.1	26.7				mg/L	12	Standard
Ho-1	165	130.0	21.4				mg/L	12	Standard
Er	166	116.7	51.7				mg/L	20	Standard
I	127	1244078.5	12.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.231	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.766	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:51:20
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.387
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	97.697
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Mo 98 Upper, S, EEE	Mo	98	
Cd 111 Lower	Cd	111	

Sample ID: L1606036203
Report Date/Time: Tuesday, June 28, 2016 13:51:20
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036204
Sample Date/Time: Tuesday, June 28, 2016 13:52:18
Number of Replicates: 3
Autosampler Position: 206
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	103725.5	0.9				ug/L	81760	Standard
	Be	9	18.3	41.7	**0.0048**	0.006	125.1	ug/L	17	Standard
	Al	27	23803073.4	2.1	**121.7007**	1.863	1.5	ug/L	1410	Standard
	Sc	45	41989.8	2.5				ug/L	39543	Standard
	Ti	47	198.3	8.2	**0.6152**	0.056	9.1	ug/L	39	Standard
	V	51	2022.8	10.2	**0.0784**	0.030	38.8	ug/L	1415	Standard
	Cr	52	8025.5	3.2	**0.0718**	0.049	68.1	ug/L	6946	Standard
	Cr	53	1725.1	14.9	**1.5845**	0.307	19.4	ug/L	328	Standard
	Mn	55	8347.3	4.1	**1.1117**	0.038	3.4	ug/L	823	Standard
	Co	59	1991.1	1.1	**0.2461**	0.003	1.1	ug/L	252	Standard
	Ni	60	2784.9	2.6	**1.3428**	0.049	3.7	ug/L	192	Standard
	Cu	65	1676.4	9.2	**0.7724**	0.072	9.4	ug/L	141	Standard
	Zn	66	3951.5	5.5	**2.7003**	0.153	5.7	ug/L	491	Standard
>	**Ge**	72	526193.8	1.1				ug/L	515087	Standard
	As	75	658.8	2.1	**0.6140**	0.018	3.0	ug/L	20	Standard
	Se	82	229.7	4.0	**1.8340**	0.092	5.0	ug/L	15	Standard
	Se-1	77	153.0	9.2	**1.2707**	0.204	16.0	ug/L	48	Standard
>	**Ga**	71	148.3	16.6				mg/L	90	Standard
	Rb	85	683.3	4.9				ug/L	23	Standard
	Y	89	453248.0	2.1				ug/L	448510	Standard
>	**Rh**	103	91.7	27.5				ug/L	15	Standard
	Mo	98	5026.0	6.3	**1.8280**	0.098	5.4	ug/L	11	Standard
	Ag	107	95.3	7.7	**-0.0021**	0.001	57.5	ug/L	134	Standard
	Cd	111	8.0	39.8	**0.0008**	0.001	178.0	mg/L	7	Standard
	Cd	114	50.5	30.0	**0.0047**	0.003	60.8	ug/L	16	Standard
>	**In**	115	615030.5	1.6				ug/L	611117	Standard
	Sn	118	1100.0	6.8	**0.0830**	0.014	16.4	ug/L	518	Standard
	Sb	123	686.5	31.6	**0.1319**	0.046	34.9	ug/L	127	Standard
	Ba	135	22769.8	0.9	**10.9347**	0.080	0.7	ug/L	45	Standard
	Ce	140	320.0	14.9				ug/L	298	Standard
>	**Tb**	159	1016190.3	1.0				ug/L	977942	Standard
	Ho	165	48.3	15.8				ug/L	12	Standard
	Tl	203	156.0	5.7	**0.0132**	0.001	7.6	ug/L	31	Standard
	Tl	205	153.3	14.7	**0.0181**	0.003	17.2	ug/L	20	Standard
	Pb	206	466.3	6.0	**0.0202**	0.005	24.4	ug/L	286	Standard
	Pb	207	405.3	5.2	**0.0227**	0.004	15.6	ug/L	257	Standard
	Pb	208	1478.0	2.2	**0.0157**	0.002	13.7	ug/L	956	Standard
	U	238	1957.8	2.0	**0.2625**	0.004	1.4	ug/L	17	Standard
>	**Bi**	209	573383.4	0.9				ug/L	557080	Standard

Sample ID: L1606036204
Report Date/Time: Tuesday, June 28, 2016 13:54:35
Page 1

Approved: June 29, 2016

Na	23	8.3	34.6	**50.1179**	16.456	32.8	mg/L	0	Standard
Mg	24	58420.8	1.2	**73.2244**	2.756	3.8	mg/L	32	Standard
K	39	723.4	8.7	**2.1601**	0.187	8.7	mg/L	30	Standard
Ca	43	463.3	11.9	**60.9213**	7.520	12.3	mg/L	48	Standard
Fe	54	191.2	7.7	**0.0260**	0.014	55.5	mg/L	159	Standard
Fe	57	531.7	2.7	**1.0656**	0.010	0.9	mg/L	195	Standard
> **Sc-1**	45	41989.8	2.5				mg/L	39543	Standard
Cl	35	37603.3	1.8				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	30841.4	4.4				ug/L	1647	Standard
P	31	21757.3	1.3				ug/L	15231	Standard
S	34	1778.4	1.5				ug/L	1845	Standard
Sr	88	246.7	9.1				ug/L	70	Standard
C	12	216.7	7.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	60.8	24.7				mg/L	12	Standard
Ho-1	165	48.3	15.8				mg/L	12	Standard
Er	166	53.3	10.8				mg/L	20	Standard
I	127	78536.8	23.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		126.866	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.156	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036204
Report Date/Time: Tuesday, June 28, 2016 13:54:35
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	100.640
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.927
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036205
Sample Date/Time: Tuesday, June 28, 2016 13:59:25
Number of Replicates: 3
Autosampler Position: 207
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	103215.4	0.9				ug/L	81760	Standard
	Be	9	20.0	50.0	**0.0063**	0.008	129.0	ug/L	17	Standard
	Al	27	14439258.9	3.3	**74.2049**	3.024	4.1	ug/L	1410	Standard
	Sc	45	39693.6	2.3				ug/L	39543	Standard
	Ti	47	140.0	4.0	**0.4180**	0.020	4.8	ug/L	39	Standard
	V	51	3604.7	5.2	**0.3305**	0.025	7.7	ug/L	1415	Standard
	Cr	52	7555.2	2.8	**0.0671**	0.025	36.8	ug/L	6946	Standard
	Cr	53	1441.7	2.8	**1.3602**	0.060	4.4	ug/L	328	Standard
	Mn	55	692335.2	1.4	**110.4896**	2.247	2.0	ug/L	823	Standard
	Co	59	1760.4	3.6	**0.2283**	0.011	4.9	ug/L	252	Standard
	Ni	60	2996.0	4.5	**1.5433**	0.085	5.5	ug/L	192	Standard
	Cu	65	1684.1	4.6	**0.8273**	0.040	4.8	ug/L	141	Standard
	Zn	66	2371.2	1.2	**1.4134**	0.020	1.4	ug/L	491	Standard
>	**Ge**	72	497232.9	0.8				ug/L	515087	Standard
	As	75	475.6	6.3	**0.4785**	0.029	6.0	ug/L	20	Standard
	Se	82	111.0	8.1	**0.8794**	0.084	9.6	ug/L	15	Standard
	Se-1	77	118.0	5.1	**0.9042**	0.095	10.5	ug/L	48	Standard
>	**Ga**	71	108.3	35.3				mg/L	90	Standard
	Rb	85	57350.0	0.9				ug/L	23	Standard
	Y	89	457066.8	2.1				ug/L	448510	Standard
>	**Rh**	103	28.3	44.4				ug/L	15	Standard
	Mo	98	255967.1	1.2	**93.6404**	2.667	2.8	ug/L	11	Standard
	Ag	107	89.3	9.3	**-0.0031**	0.001	44.5	ug/L	134	Standard
	Cd	111	-197.3	3.0	**-0.0942**	0.001	1.4	mg/L	7	Standard
	Cd	114	425.0	6.1	**0.0760**	0.006	8.0	ug/L	16	Standard
>	**In**	115	618093.1	1.7				ug/L	611117	Standard
	Sn	118	871.7	7.6	**0.0437**	0.009	20.3	ug/L	518	Standard
	Sb	123	474.5	27.0	**0.0866**	0.026	30.6	ug/L	127	Standard
	Ba	135	13314.9	2.5	**6.3544**	0.263	4.1	ug/L	45	Standard
	Ce	140	228.3	6.3				ug/L	298	Standard
>	**Tb**	159	1010942.5	1.3				ug/L	977942	Standard
	Ho	165	25.0	72.1				ug/L	12	Standard
	Tl	203	239.7	6.5	**0.0241**	0.002	9.5	ug/L	31	Standard
	Tl	205	181.7	22.1	**0.0230**	0.006	26.1	ug/L	20	Standard
	Pb	206	319.7	6.3	**-0.0035**	0.004	108.4	ug/L	286	Standard
	Pb	207	282.7	2.0	**0.0009**	0.002	252.3	ug/L	257	Standard
	Pb	208	1120.3	3.0	**-0.0001**	0.003	3300.1	ug/L	956	Standard
	U	238	733.7	3.0	**0.1011**	0.005	4.9	ug/L	17	Standard
>	**Bi**	209	548318.3	1.8				ug/L	557080	Standard

Sample ID: L1606036205
Report Date/Time: Tuesday, June 28, 2016 14:01:41
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**21.6860**	37.553	173.2	mg/L	0	Standard
Mg	24	97569.9	1.9	**129.3694**	3.560	2.8	mg/L	32	Standard
K	39	4389.0	6.2	**14.6373**	1.001	6.8	mg/L	30	Standard
Ca	43	378.3	2.8	**51.9377**	2.720	5.2	mg/L	48	Standard
Fe	54	155.8	7.4	**0.0035**	0.012	336.9	mg/L	159	Standard
Fe	57	446.7	11.3	**0.8756**	0.135	15.4	mg/L	195	Standard
> **Sc-1**	45	39693.6	2.3				mg/L	39543	Standard
Cl	35	45044.9	1.9				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	18554.7	4.8				ug/L	1647	Standard
P	31	20283.6	3.2				ug/L	15231	Standard
S	34	1950.1	2.9				ug/L	1845	Standard
Sr	88	140.0	14.3				ug/L	70	Standard
C	12	253.3	26.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	47.4	35.9				mg/L	12	Standard
Ho-1	165	25.0	72.1				mg/L	12	Standard
Er	166	53.3	21.7				mg/L	20	Standard
I	127	81681.7	5.6				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		126.242	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		96.534	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606036205
Report Date/Time: Tuesday, June 28, 2016 14:01:41
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	101.142	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	98.427	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036206
Sample Date/Time: Tuesday, June 28, 2016 14:02:36
Number of Replicates: 3
Autosampler Position: 208
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	108188.3	1.9				ug/L	81760	Standard
	Be	9	25.0	69.3	**0.0094**	0.014	144.0	ug/L	17	Standard
	Al	27	23235387.8	1.2	**113.9176**	1.672	1.5	ug/L	1410	Standard
	Sc	45	42060.0	3.5				ug/L	39543	Standard
	Ti	47	152.7	7.7	**0.4364**	0.055	12.6	ug/L	39	Standard
	V	51	2674.6	4.0	**0.1700**	0.012	6.8	ug/L	1415	Standard
	Cr	52	8527.1	1.1	**0.1424**	0.031	21.5	ug/L	6946	Standard
	Cr	53	1513.4	3.0	**1.3470**	0.073	5.5	ug/L	328	Standard
	Mn	55	17845.5	5.4	**2.5461**	0.135	5.3	ug/L	823	Standard
	Co	59	1119.4	15.5	**0.1245**	0.023	18.8	ug/L	252	Standard
	Ni	60	2702.9	1.5	**1.3004**	0.041	3.1	ug/L	192	Standard
	Cu	65	753.4	4.7	**0.2972**	0.013	4.4	ug/L	141	Standard
	Zn	66	1481.1	2.0	**0.4954**	0.035	7.0	ug/L	491	Standard
>	**Ge**	72	526170.8	1.4				ug/L	515087	Standard
	As	75	291.6	5.2	**0.2943**	0.014	4.7	ug/L	20	Standard
	Se	82	87.6	19.6	**0.6269**	0.155	24.7	ug/L	15	Standard
	Se-1	77	124.0	13.8	**0.8941**	0.239	26.7	ug/L	48	Standard
>	**Ga**	71	93.3	13.5				mg/L	90	Standard
	Rb	85	1890.1	1.5				ug/L	23	Standard
	Y	89	457488.4	1.5				ug/L	448510	Standard
>	**Rh**	103	41.7	45.4				ug/L	15	Standard
	Mo	98	1289.7	27.8	**0.4294**	0.125	29.1	ug/L	11	Standard
	Ag	107	132.0	62.6	**0.0022**	0.011	517.7	ug/L	134	Standard
	Cd	111	14.8	9.0	**0.0036**	0.001	18.2	mg/L	7	Standard
	Cd	114	35.3	64.8	**0.0015**	0.004	279.2	ug/L	16	Standard
>	**In**	115	650440.2	1.2				ug/L	611117	Standard
	Sn	118	795.0	5.8	**0.0244**	0.006	26.1	ug/L	518	Standard
	Sb	123	299.9	40.5	**0.0467**	0.023	50.3	ug/L	127	Standard
	Ba	135	16583.1	1.2	**7.5223**	0.109	1.5	ug/L	45	Standard
	Ce	140	573.3	12.7				ug/L	298	Standard
>	**Tb**	159	1070413.0	0.5				ug/L	977942	Standard
	Ho	165	35.0	14.3				ug/L	12	Standard
	Tl	203	143.0	68.4	**0.0109**	0.011	100.1	ug/L	31	Standard
	Tl	205	173.3	55.8	**0.0194**	0.012	61.7	ug/L	20	Standard
	Pb	206	407.0	5.8	**0.0055**	0.005	85.2	ug/L	286	Standard
	Pb	207	361.0	20.9	**0.0099**	0.014	143.4	ug/L	257	Standard
	Pb	208	1551.4	24.3	**0.0150**	0.018	119.8	ug/L	956	Standard
	U	238	2137.8	4.4	**0.2708**	0.011	4.2	ug/L	17	Standard
>	**Bi**	209	607194.4	1.0				ug/L	557080	Standard

Sample ID: L1606036206
Report Date/Time: Tuesday, June 28, 2016 14:05:08
Page 1

Approved: June 29, 2016

Na	23	15.0	57.7	**91.7381**	54.279	59.2	mg/L	0	Standard
Mg	24	7570.2	0.6	**9.4255**	0.372	3.9	mg/L	32	Standard
K	39	228.3	12.1	**0.5833**	0.074	12.7	mg/L	30	Standard
Ca	43	403.3	9.3	**52.3767**	6.757	12.9	mg/L	48	Standard
Fe	54	189.4	6.8	**0.0240**	0.007	29.2	mg/L	159	Standard
Fe	57	486.7	15.4	**0.9208**	0.241	26.2	mg/L	195	Standard
> **Sc-1**	45	42060.0	3.5				mg/L	39543	Standard
Cl	35	35390.6	1.2				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	8392.3	5.9				ug/L	1647	Standard
P	31	21440.2	3.3				ug/L	15231	Standard
S	34	2026.8	8.3				ug/L	1845	Standard
Sr	88	143.3	29.0				ug/L	70	Standard
C	12	223.3	9.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	47.9	21.4				mg/L	12	Standard
Ho-1	165	35.0	14.3				mg/L	12	Standard
Er	166	43.3	35.3				mg/L	20	Standard
I	127	10155.1	10.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.324	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.152	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036206
Report Date/Time: Tuesday, June 28, 2016 14:05:08
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	106.435
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.996
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036207
Sample Date/Time: Tuesday, June 28, 2016 14:06:02
Number of Replicates: 3
Autosampler Position: 209
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	107639.4	1.9				ug/L	81760	Standard
	Be	9	20.0	0.0	**0.0056**	0.000	5.4	ug/L	17	Standard
	Al	27	3476504.2	1.7	**17.1173**	0.043	0.3	ug/L	1410	Standard
	Sc	45	39367.8	2.7				ug/L	39543	Standard
	Ti	47	100.3	3.5	**0.2518**	0.007	3.0	ug/L	39	Standard
	V	51	5441.1	3.8	**0.6024**	0.020	3.4	ug/L	1415	Standard
	Cr	52	7747.6	3.6	**0.0931**	0.026	27.8	ug/L	6946	Standard
	Cr	53	1961.8	9.7	**1.9650**	0.153	7.8	ug/L	328	Standard
	Mn	55	12138.9	1.5	**1.7871**	0.034	1.9	ug/L	823	Standard
	Co	59	2054.5	1.6	**0.2711**	0.013	4.9	ug/L	252	Standard
	Ni	60	1564.4	4.1	**0.7535**	0.028	3.7	ug/L	192	Standard
	Cu	65	3547.1	3.0	**1.8404**	0.110	6.0	ug/L	141	Standard
	Zn	66	3174.7	2.0	**2.1684**	0.146	6.7	ug/L	491	Standard
>	**Ge**	72	498440.9	3.0				ug/L	515087	Standard
	As	75	2063.4	4.9	**1.9397**	0.147	7.6	ug/L	20	Standard
	Se	82	279.9	8.8	**2.3986**	0.288	12.0	ug/L	15	Standard
	Se-1	77	196.7	12.4	**1.9867**	0.408	20.5	ug/L	48	Standard
>	**Ga**	71	203.3	22.3				mg/L	90	Standard
	Rb	85	4590.7	1.6				ug/L	23	Standard
	Y	89	429873.2	1.1				ug/L	448510	Standard
>	**Rh**	103	36.7	15.7				ug/L	15	Standard
	Mo	98	411573.4	1.6	**157.7590**	2.905	1.8	ug/L	11	Standard
	Ag	107	97.0	12.7	**-0.0013**	0.002	144.7	ug/L	134	Standard
	Cd	111	-313.3	8.5	**-0.1547**	0.012	7.7	mg/L	7	Standard
	Cd	114	654.0	5.6	**0.1255**	0.007	5.6	ug/L	16	Standard
>	**In**	115	589803.8	0.9				ug/L	611117	Standard
	Sn	118	783.4	15.3	**0.0353**	0.020	57.5	ug/L	518	Standard
	Sb	123	727.4	9.3	**0.1468**	0.014	9.3	ug/L	127	Standard
	Ba	135	15565.3	1.0	**7.7872**	0.095	1.2	ug/L	45	Standard
	Ce	140	646.7	2.7				ug/L	298	Standard
>	**Tb**	159	991896.0	2.4				ug/L	977942	Standard
	Ho	165	86.7	18.5				ug/L	12	Standard
	Tl	203	276.7	3.3	**0.0282**	0.002	5.4	ug/L	31	Standard
	Tl	205	283.3	3.7	**0.0365**	0.002	4.5	ug/L	20	Standard
	Pb	206	399.0	5.4	**0.0104**	0.005	44.3	ug/L	286	Standard
	Pb	207	325.3	5.9	**0.0089**	0.004	45.8	ug/L	257	Standard
	Pb	208	1296.0	0.3	**0.0084**	0.001	12.8	ug/L	956	Standard
	U	238	1382.7	4.5	**0.1905**	0.008	4.1	ug/L	17	Standard
>	**Bi**	209	555677.7	1.4				ug/L	557080	Standard

Sample ID: L1606036207
Report Date/Time: Tuesday, June 28, 2016 14:08:19
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**21.6985**	18.802	86.6	mg/L	0	Standard
Mg	24	201508.1	2.0	**269.4503**	5.984	2.2	mg/L	32	Standard
K	39	1843.4	2.4	**6.1200**	0.302	4.9	mg/L	30	Standard
Ca	43	86.7	16.7	**7.9004**	2.376	30.1	mg/L	48	Standard
Fe	54	134.1	6.8	**-0.0152**	0.007	45.6	mg/L	159	Standard
Fe	57	218.3	31.3	**0.1119**	0.215	192.0	mg/L	195	Standard
> **Sc-1**	45	39367.8	2.7				mg/L	39543	Standard
Cl	35	57177.0	0.5				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	35885.9	6.7				ug/L	1647	Standard
P	31	17845.5	1.4				ug/L	15231	Standard
S	34	1995.1	3.9				ug/L	1845	Standard
Sr	88	96.7	26.5				ug/L	70	Standard
C	12	250.0	28.0				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	68.0	19.9				mg/L	12	Standard
Ho-1	165	86.7	18.5				mg/L	12	Standard
Er	166	180.0	33.3				mg/L	20	Standard
I	127	352295.0	10.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		131.653	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.768	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036207
Report Date/Time: Tuesday, June 28, 2016 14:08:19
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.512
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.748
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mo 98 Upper, S, EEE	Mo	98	
Cd 111 Lower	Cd	111	

Sample ID: L1606036207
Report Date/Time: Tuesday, June 28, 2016 14:08:19
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036208
Sample Date/Time: Tuesday, June 28, 2016 14:09:14
Number of Replicates: 3
Autosampler Position: 210
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	113878.0	0.7				ug/L	81760	Standard
	Be	9	15.0	33.3	**0.0011**	0.004	325.9	ug/L	17	Standard
	Al	27	36051659.4	1.3	**167.9026**	1.365	0.8	ug/L	1410	Standard
	Sc	45	40123.1	2.7				ug/L	39543	Standard
	Ti	47	157.7	1.6	**0.4795**	0.006	1.3	ug/L	39	Standard
	V	51	2374.4	7.5	**0.1409**	0.028	20.1	ug/L	1415	Standard
	Cr	52	8200.6	0.9	**0.1408**	0.017	11.9	ug/L	6946	Standard
	Cr	53	1531.7	7.2	**1.4329**	0.130	9.1	ug/L	328	Standard
	Mn	55	2760.9	2.4	**0.2845**	0.009	3.3	ug/L	823	Standard
	Co	59	1114.0	4.4	**0.1299**	0.007	5.8	ug/L	252	Standard
	Ni	60	3861.2	1.0	**1.9798**	0.033	1.7	ug/L	192	Standard
	Cu	65	1286.1	5.6	**0.5974**	0.039	6.5	ug/L	141	Standard
	Zn	66	2133.2	3.4	**1.1513**	0.069	6.0	ug/L	491	Standard
>	**Ge**	72	506573.0	0.6				ug/L	515087	Standard
	As	75	124.6	44.5	**0.1528**	0.050	32.6	ug/L	20	Standard
	Se	82	87.7	7.6	**0.6552**	0.055	8.5	ug/L	15	Standard
	Se-1	77	126.7	6.6	**0.9909**	0.116	11.7	ug/L	48	Standard
>	**Ga**	71	56.7	45.3				mg/L	90	Standard
	Rb	85	275.0	18.5				ug/L	23	Standard
	Y	89	450902.0	1.5				ug/L	448510	Standard
>	**Rh**	103	46.7	32.7				ug/L	15	Standard
	Mo	98	3181.2	4.8	**1.1379**	0.054	4.7	ug/L	11	Standard
	Ag	107	86.0	6.5	**-0.0036**	0.001	24.6	ug/L	134	Standard
	Cd	111	11.8	23.2	**0.0025**	0.001	50.9	mg/L	7	Standard
	Cd	114	54.8	27.0	**0.0054**	0.003	50.7	ug/L	16	Standard
>	**In**	115	621612.8	0.6				ug/L	611117	Standard
	Sn	118	878.4	13.7	**0.0442**	0.021	46.4	ug/L	518	Standard
	Sb	123	218.7	21.6	**0.0327**	0.010	30.0	ug/L	127	Standard
	Ba	135	24511.5	2.2	**11.6480**	0.297	2.5	ug/L	45	Standard
	Ce	140	126.7	17.8				ug/L	298	Standard
>	**Tb**	159	1020349.4	1.6				ug/L	977942	Standard
	Ho	165	38.3	41.9				ug/L	12	Standard
	Tl	203	70.7	5.4	**0.0034**	0.000	14.1	ug/L	31	Standard
	Tl	205	50.0	36.1	**0.0044**	0.002	54.4	ug/L	20	Standard
	Pb	206	339.3	6.0	**-0.0025**	0.003	133.5	ug/L	286	Standard
	Pb	207	287.3	3.0	**-0.0007**	0.002	279.9	ug/L	257	Standard
	Pb	208	1133.4	0.9	**-0.0020**	0.001	40.2	ug/L	956	Standard
	U	238	2248.5	2.1	**0.3023**	0.005	1.6	ug/L	17	Standard
>	**Bi**	209	572724.5	0.5				ug/L	557080	Standard

Sample ID: L1606036208
Report Date/Time: Tuesday, June 28, 2016 14:11:31
Page 1

Approved: June 29, 2016

Na	23	11.7	65.5	**74.1428**	49.752	67.1	mg/L	0	Standard
Mg	24	18449.6	1.2	**24.1624**	0.882	3.7	mg/L	32	Standard
K	39	318.3	12.8	**0.9203**	0.151	16.4	mg/L	30	Standard
Ca	43	671.7	8.0	**95.1609**	6.901	7.3	mg/L	48	Standard
Fe	54	172.3	21.8	**0.0165**	0.033	197.4	mg/L	159	Standard
Fe	57	548.3	9.0	**1.2014**	0.184	15.3	mg/L	195	Standard
> **Sc-1**	45	40123.1	2.7				mg/L	39543	Standard
Cl	35	45031.5	1.7				ug/L	27642	Standard
Kr	83	2.3	89.2				ug/L	2	Standard
Br	81	9216.1	3.8				ug/L	1647	Standard
P	31	22331.4	1.4				ug/L	15231	Standard
S	34	2088.5	6.2				ug/L	1845	Standard
Sr	88	246.7	15.5				ug/L	70	Standard
C	12	346.7	6.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	37.4	4.1				mg/L	12	Standard
Ho-1	165	38.3	41.9				mg/L	12	Standard
Er	166	53.3	60.3				mg/L	20	Standard
I	127	16496.5	38.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		139.283	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.347	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036208
Report Date/Time: Tuesday, June 28, 2016 14:11:31
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.718
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.808
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036209
Sample Date/Time: Tuesday, June 28, 2016 14:12:25
Number of Replicates: 3
Autosampler Position: 211
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	107575.7	2.1				ug/L	81760	Standard
	Be	9	15.0	33.3	**0.0018**	0.004	223.3	ug/L	17	Standard
	Al	27	24368759.8	1.3	**120.1696**	2.653	2.2	ug/L	1410	Standard
	Sc	45	40144.8	3.7				ug/L	39543	Standard
	Ti	47	175.7	9.7	**0.5394**	0.064	11.9	ug/L	39	Standard
	V	51	1717.1	6.1	**0.0398**	0.013	33.2	ug/L	1415	Standard
	Cr	52	8804.6	0.4	**0.2048**	0.007	3.4	ug/L	6946	Standard
	Cr	53	1576.7	7.9	**1.4512**	0.154	10.6	ug/L	328	Standard
	Mn	55	1083613.1	0.9	**166.6059**	0.333	0.2	ug/L	823	Standard
	Co	59	2591.6	2.6	**0.3366**	0.007	2.1	ug/L	252	Standard
	Ni	60	4161.2	1.9	**2.0993**	0.030	1.4	ug/L	192	Standard
	Cu	65	1088.7	1.4	**0.4808**	0.011	2.3	ug/L	141	Standard
	Zn	66	3226.7	2.3	**2.1094**	0.068	3.2	ug/L	491	Standard
>	**Ge**	72	516304.1	0.8				ug/L	515087	Standard
	As	75	854.1	2.3	**0.7983**	0.019	2.4	ug/L	20	Standard
	Se	82	52.5	11.5	**0.3355**	0.056	16.6	ug/L	15	Standard
	Se-1	77	105.7	6.3	**0.6809**	0.097	14.3	ug/L	48	Standard
>	**Ga**	71	153.3	30.3				mg/L	90	Standard
	Rb	85	15286.0	2.4				ug/L	23	Standard
	Y	89	449805.6	1.9				ug/L	448510	Standard
>	**Rh**	103	36.7	43.8				ug/L	15	Standard
	Mo	98	1903.1	2.7	**0.6696**	0.018	2.6	ug/L	11	Standard
	Ag	107	88.0	6.3	**-0.0034**	0.001	30.6	ug/L	134	Standard
	Cd	111	24.4	20.6	**0.0083**	0.002	29.8	mg/L	7	Standard
	Cd	114	75.3	47.3	**0.0092**	0.007	71.3	ug/L	16	Standard
>	**In**	115	624865.1	1.8				ug/L	611117	Standard
	Sn	118	858.4	10.9	**0.0403**	0.018	44.8	ug/L	518	Standard
	Sb	123	598.6	8.7	**0.1114**	0.013	11.4	ug/L	127	Standard
	Ba	135	79235.0	1.1	**37.5150**	0.511	1.4	ug/L	45	Standard
	Ce	140	1043.4	5.0				ug/L	298	Standard
>	**Tb**	159	1006271.6	1.0				ug/L	977942	Standard
	Ho	165	31.7	55.5				ug/L	12	Standard
	Tl	203	94.7	10.4	**0.0060**	0.001	17.2	ug/L	31	Standard
	Tl	205	103.3	26.6	**0.0113**	0.003	31.1	ug/L	20	Standard
	Pb	206	1317.1	2.4	**0.1704**	0.008	4.5	ug/L	286	Standard
	Pb	207	1098.7	2.3	**0.1583**	0.006	3.8	ug/L	257	Standard
	Pb	208	4505.3	1.8	**0.1689**	0.006	3.3	ug/L	956	Standard
	U	238	1688.4	2.2	**0.2232**	0.003	1.5	ug/L	17	Standard
>	**Bi**	209	580367.6	0.9				ug/L	557080	Standard

Sample ID: L1606036209
Report Date/Time: Tuesday, June 28, 2016 14:14:42
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**32.4402**	32.912	101.5	mg/L	0	Standard
Mg	24	8242.2	2.3	**10.7549**	0.236	2.2	mg/L	32	Standard
K	39	375.0	23.7	**1.1102**	0.318	28.6	mg/L	30	Standard
Ca	43	428.3	18.6	**58.5393**	9.751	16.7	mg/L	48	Standard
Fe	54	1288.6	9.8	**1.0145**	0.141	13.9	mg/L	159	Standard
Fe	57	713.4	8.9	**1.7461**	0.160	9.2	mg/L	195	Standard
> **Sc-1**	45	40144.8	3.7				mg/L	39543	Standard
Cl	35	37853.9	1.3				ug/L	27642	Standard
Kr	83	2.3	173.2				ug/L	2	Standard
Br	81	8949.3	3.3				ug/L	1647	Standard
P	31	22830.5	2.4				ug/L	15231	Standard
S	34	2348.5	6.0				ug/L	1845	Standard
Sr	88	131.7	18.7				ug/L	70	Standard
C	12	320.0	23.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	57.1	19.2				mg/L	12	Standard
Ho-1	165	31.7	55.5				mg/L	12	Standard
Er	166	60.0	44.1				mg/L	20	Standard
I	127	12798.7	0.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		131.575	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.236	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036209
Report Date/Time: Tuesday, June 28, 2016 14:14:42
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	102.250
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.180
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606036209
Report Date/Time: Tuesday, June 28, 2016 14:14:42
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036210
Sample Date/Time: Tuesday, June 28, 2016 14:15:37
Number of Replicates: 3
Autosampler Position: 212
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	112279.5	1.4				ug/L	81760	Standard
	Be	9	23.3	32.7	**0.0074**	0.006	75.4	ug/L	17	Standard
	Al	27	11519588.8	0.5	**54.4098**	0.542	1.0	ug/L	1410	Standard
	Sc	45	38345.1	1.2				ug/L	39543	Standard
	Ti	47	94.7	7.9	**0.2480**	0.026	10.6	ug/L	39	Standard
	V	51	2217.7	3.9	**0.1400**	0.014	10.3	ug/L	1415	Standard
	Cr	52	7580.6	0.1	**0.1258**	0.020	16.0	ug/L	6946	Standard
	Cr	53	2436.9	5.1	**2.6754**	0.201	7.5	ug/L	328	Standard
	Mn	55	75388.8	2.2	**12.4826**	0.333	2.7	ug/L	823	Standard
	Co	59	4144.2	3.9	**0.6094**	0.028	4.6	ug/L	252	Standard
	Ni	60	4832.4	1.6	**2.6820**	0.049	1.8	ug/L	192	Standard
	Cu	65	4625.4	2.6	**2.5532**	0.075	2.9	ug/L	141	Standard
	Zn	66	4203.9	2.1	**3.3383**	0.098	2.9	ug/L	491	Standard
>	**Ge**	72	474293.4	1.6				ug/L	515087	Standard
	As	75	1073.0	2.0	**1.0772**	0.038	3.5	ug/L	20	Standard
	Se	82	351.4	3.1	**3.1960**	0.130	4.1	ug/L	15	Standard
	Se-1	77	210.7	7.1	**2.3175**	0.198	8.5	ug/L	48	Standard
>	**Ga**	71	73.3	33.6				mg/L	90	Standard
	Rb	85	56526.9	1.1				ug/L	23	Standard
	Y	89	421082.0	1.6				ug/L	448510	Standard
>	**Rh**	103	96.7	14.9				ug/L	15	Standard
	Mo	98	257904.5	1.0	**100.5156**	1.205	1.2	ug/L	11	Standard
	Ag	107	79.7	12.8	**-0.0037**	0.002	45.1	ug/L	134	Standard
	Cd	111	-189.4	16.3	**-0.0962**	0.015	15.7	mg/L	7	Standard
	Cd	114	358.5	16.9	**0.0677**	0.012	17.5	ug/L	16	Standard
>	**In**	115	580013.8	0.6				ug/L	611117	Standard
	Sn	118	865.0	5.3	**0.0522**	0.009	16.9	ug/L	518	Standard
	Sb	123	234.3	5.5	**0.0395**	0.003	7.3	ug/L	127	Standard
	Ba	135	12298.0	0.4	**6.2513**	0.035	0.6	ug/L	45	Standard
	Ce	140	386.7	8.8				ug/L	298	Standard
>	**Tb**	159	979000.4	0.2				ug/L	977942	Standard
	Ho	165	86.7	26.6				ug/L	12	Standard
	Tl	203	556.3	8.3	**0.0635**	0.006	10.1	ug/L	31	Standard
	Tl	205	491.7	8.6	**0.0672**	0.007	10.1	ug/L	20	Standard
	Pb	206	339.7	4.4	**0.0013**	0.002	176.0	ug/L	286	Standard
	Pb	207	303.3	2.7	**0.0062**	0.001	17.9	ug/L	257	Standard
	Pb	208	1090.3	1.9	**-0.0007**	0.002	299.6	ug/L	956	Standard
	U	238	117.3	12.8	**0.0140**	0.002	17.0	ug/L	17	Standard
>	**Bi**	209	539461.3	1.8				ug/L	557080	Standard

Sample ID: L1606036210
Report Date/Time: Tuesday, June 28, 2016 14:17:54
Page 1

Approved: June 29, 2016

Na	23	8.3	34.6	**55.0323**	18.599	33.8	mg/L	0	Standard
Mg	24	255159.1	1.2	**350.2417**	6.027	1.7	mg/L	32	Standard
K	39	5237.6	3.5	**18.1047**	0.448	2.5	mg/L	30	Standard
Ca	43	815.0	11.2	**122.4307**	15.864	13.0	mg/L	48	Standard
Fe	54	150.7	21.1	**0.0038**	0.031	813.1	mg/L	159	Standard
Fe	57	818.4	11.3	**2.2227**	0.299	13.5	mg/L	195	Standard
> **Sc-1**	45	38345.1	1.2				mg/L	39543	Standard
Cl	35	84049.9	1.6				ug/L	27642	Standard
Kr	83	2.0	100.0				ug/L	2	Standard
Br	81	57144.3	3.2				ug/L	1647	Standard
P	31	18494.6	4.2				ug/L	15231	Standard
S	34	1918.5	8.3				ug/L	1845	Standard
Sr	88	445.0	8.9				ug/L	70	Standard
C	12	253.3	16.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	64.7	80.3				mg/L	12	Standard
Ho-1	165	86.7	26.6				mg/L	12	Standard
Er	166	110.0	9.1				mg/L	20	Standard
I	127	999438.0	9.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		137.328	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.080	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036210
Report Date/Time: Tuesday, June 28, 2016 14:17:54
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.911
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.837
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mo 98 Upper, S, EEE	Mo	98	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 14:18:50
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	108145.2	5.1				ug/L	81760	Standard
	Be	9	64242.8	4.5	**48.9113**	0.992	2.0	ug/L	17	Standard
	Al	27	8304059.2	3.3	**40.7429**	1.039	2.5	ug/L	1410	Standard
	Sc	45	37765.4	4.0				ug/L	39543	Standard
	Ti	47	23471.8	2.0	**97.1021**	1.269	1.3	ug/L	39	Standard
	V	51	349107.5	0.6	**51.5604**	0.772	1.5	ug/L	1415	Standard
	Cr	52	335198.9	1.2	**48.5745**	0.974	2.0	ug/L	6946	Standard
	Cr	53	41869.4	1.3	**48.8267**	1.454	3.0	ug/L	328	Standard
	Mn	55	314993.6	1.3	**49.9295**	0.954	1.9	ug/L	823	Standard
	Co	59	375345.6	0.1	**55.0339**	1.228	2.2	ug/L	252	Standard
	Ni	60	88766.6	0.5	**48.4746**	0.838	1.7	ug/L	192	Standard
	Cu	65	90675.9	1.2	**49.0916**	0.682	1.4	ug/L	141	Standard
	Zn	66	53067.8	0.6	**49.0759**	0.955	1.9	ug/L	491	Standard
>	Ge	72	499870.2	2.1				ug/L	515087	Standard
	As	75	55072.1	0.8	**50.5344**	1.449	2.9	ug/L	20	Standard
	Se	82	6022.1	1.2	**53.7900**	0.803	1.5	ug/L	15	Standard
	Se-1	77	3835.5	2.7	**51.7480**	1.296	2.5	ug/L	48	Standard
>	Ga	71	68.3	34.6				mg/L	90	Standard
	Rb	85	878.4	4.0				ug/L	23	Standard
	Y	89	444195.9	1.0				ug/L	448510	Standard
>	Rh	103	36.7	7.9				ug/L	15	Standard
	Mo	98	271765.3	0.4	**101.0586**	1.139	1.1	ug/L	11	Standard
	Ag	107	335937.1	0.8	**50.2935**	0.228	0.5	ug/L	134	Standard
	Cd	111	104881.4	1.9	**49.3201**	0.733	1.5	mg/L	7	Standard
	Cd	114	258223.7	2.1	**49.9610**	1.276	2.6	ug/L	16	Standard
>	In	115	607920.9	0.8				ug/L	611117	Standard
	Sn	118	297159.7	1.8	**50.4160**	0.583	1.2	ug/L	518	Standard
	Sb	123	233566.6	0.5	**49.7525**	0.427	0.9	ug/L	127	Standard
	Ba	135	100029.2	1.5	**48.6806**	0.509	1.0	ug/L	45	Standard
	Ce	140	56.7	27.0				ug/L	298	Standard
>	Tb	159	990459.3	1.9				ug/L	977942	Standard
	Ho	165	13.3	21.7				ug/L	12	Standard
	Tl	203	438180.8	0.9	**49.9571**	0.368	0.7	ug/L	31	Standard
	Tl	205	400912.4	1.6	**52.6337**	0.214	0.4	ug/L	20	Standard
	Pb	206	285968.6	0.8	**50.7733**	0.808	1.6	ug/L	286	Standard
	Pb	207	258378.5	0.4	**50.8616**	0.645	1.3	ug/L	257	Standard
	Pb	208	1010283.1	0.5	**51.4021**	0.498	1.0	ug/L	956	Standard
	U	238	389330.8	1.6	**52.1963**	0.349	0.7	ug/L	17	Standard
>	Bi	209	579902.5	1.4				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:21:06
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4220.6	3.3	**5.8326**	0.328	5.6	mg/L	32	Standard
K	39	1343.4	3.1	**4.6154**	0.216	4.7	mg/L	30	Standard
Ca	43	66.7	37.7	**5.1699**	3.623	70.1	mg/L	48	Standard
Fe	54	5699.2	5.0	**5.2644**	0.205	3.9	mg/L	159	Standard
Fe	57	1620.1	3.2	**5.1052**	0.186	3.6	mg/L	195	Standard
> **Sc-1**	45	37765.4	4.0				mg/L	39543	Standard
Cl	35	31900.9	3.3				ug/L	27642	Standard
Kr	83	2.3	89.2				ug/L	2	Standard
Br	81	2500.2	16.5				ug/L	1647	Standard
P	31	18272.7	3.1				ug/L	15231	Standard
S	34	2018.5	1.5				ug/L	1845	Standard
Sr	88	43.3	24.0				ug/L	70	Standard
C	12	196.7	21.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	12.4	50.3				mg/L	12	Standard
Ho-1	165	13.3	21.7				mg/L	12	Standard
Er	166	20.0	86.6				mg/L	20	Standard
I	127	32500.9	63.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	97.823		
Al	27	81.486		
Sc	45			
Ti	47	97.102		
V	51	103.121		
Cr	52	97.149		
Cr	53			
Mn	55	99.859		
Co	59	110.068		
Ni	60	96.949		
Cu	65	98.183		
Zn	66	98.152		
> Ge	72		97.046	
As	75	101.069		
Se	82	107.580		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:21:06
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.059		
Ag	107	100.587		
Cd	111	98.640		
Cd	114			
In	115		99.477	
Sn	118	100.832		
Sb	123	99.505		
Ba	135	97.361		
Ce	140			
Tb	159			
Ho	165			
Tl	203	99.914		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.804		
U	238	104.393		
Bi	209		104.097	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:21:06
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 14:22:01
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101364.8	2.9				ug/L	81760	Standard
	Be	9	38.3	49.4	**0.0216**	0.016	74.7	ug/L	17	Standard
	Al	27	15489.3	94.8	**0.0675**	0.079	117.4	ug/L	1410	Standard
	Sc	45	38064.4	0.9				ug/L	39543	Standard
	Ti	47	37.3	29.4	**-0.0111**	0.045	403.8	ug/L	39	Standard
	V	51	1161.3	1.6	**-0.0351**	0.002	6.4	ug/L	1415	Standard
	Cr	52	5982.8	2.5	**-0.1747**	0.020	11.4	ug/L	6946	Standard
	Cr	53	318.3	7.1	**0.0308**	0.025	81.8	ug/L	328	Standard
	Mn	55	894.7	9.5	**-0.0068**	0.013	197.2	ug/L	823	Standard
	Co	59	241.0	14.9	**0.0039**	0.005	138.8	ug/L	252	Standard
	Ni	60	266.0	10.4	**0.0400**	0.015	37.4	ug/L	192	Standard
	Cu	65	200.0	8.9	**0.0172**	0.009	54.5	ug/L	141	Standard
	Zn	66	430.0	7.9	**-0.4249**	0.030	7.1	ug/L	491	Standard
>	**Ge**	72	501700.0	0.4				ug/L	515087	Standard
	As	75	1.5	1379.3	**0.0417**	0.019	44.5	ug/L	20	Standard
	Se	82	27.3	19.7	**0.1242**	0.049	39.3	ug/L	15	Standard
	Se-1	77	54.3	18.4	**0.0219**	0.139	635.7	ug/L	48	Standard
>	**Ga**	71	63.3	9.1				mg/L	90	Standard
	Rb	85	26.7	28.6				ug/L	23	Standard
	Y	89	434979.7	1.5				ug/L	448510	Standard
>	**Rh**	103	11.7	89.2				ug/L	15	Standard
	Mo	98	412.1	20.1	**0.1325**	0.026	19.5	ug/L	11	Standard
	Ag	107	133.7	24.7	**0.0036**	0.005	131.9	ug/L	134	Standard
	Cd	111	17.6	36.9	**0.0053**	0.003	55.6	mg/L	7	Standard
	Cd	114	37.0	22.5	**0.0022**	0.002	81.0	ug/L	16	Standard
>	**In**	115	612643.0	3.0				ug/L	611117	Standard
	Sn	118	796.7	19.5	**0.0328**	0.028	85.4	ug/L	518	Standard
	Sb	123	465.9	23.7	**0.0860**	0.025	29.2	ug/L	127	Standard
	Ba	135	55.7	25.4	**0.0018**	0.007	380.6	ug/L	45	Standard
	Ce	140	23.3	53.9				ug/L	298	Standard
>	**Tb**	159	991657.2	2.1				ug/L	977942	Standard
	Ho	165	6.7	114.6				ug/L	12	Standard
	Tl	203	81.3	30.4	**0.0043**	0.003	63.5	ug/L	31	Standard
	Tl	205	65.0	13.3	**0.0061**	0.001	17.5	ug/L	20	Standard
	Pb	206	313.0	2.4	**-0.0091**	0.002	17.1	ug/L	286	Standard
	Pb	207	268.3	2.3	**-0.0063**	0.001	16.7	ug/L	257	Standard
	Pb	208	1057.3	6.3	**-0.0078**	0.003	38.5	ug/L	956	Standard
	U	238	61.3	40.7	**0.0051**	0.003	63.9	ug/L	17	Standard
>	**Bi**	209	592397.8	0.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 14:24:18
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	95.0	63.8	**0.0743**	0.084	112.6	mg/L	32	Standard
K	39	16.7	17.3	**-0.0829**	0.011	12.9	mg/L	30	Standard
Ca	43	31.7	36.5	**-0.3389**	1.874	552.9	mg/L	48	Standard
Fe	54	110.8	32.6	**-0.0331**	0.033	100.8	mg/L	159	Standard
Fe	57	168.3	28.5	**-0.0341**	0.169	494.5	mg/L	195	Standard
> **Sc-1**	45	38064.4	0.9				mg/L	39543	Standard
Cl	35	31593.6	0.5				ug/L	27642	Standard
Kr	83	0.3	173.2				ug/L	2	Standard
Br	81	1856.8	9.2				ug/L	1647	Standard
P	31	17672.0	1.1				ug/L	15231	Standard
S	34	2055.1	6.3				ug/L	1845	Standard
Sr	88	56.7	18.4				ug/L	70	Standard
C	12	170.0	20.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	6.3	173.2				mg/L	12	Standard
Ho-1	165	6.7	114.6				mg/L	12	Standard
Er	166	6.7	173.2				mg/L	20	Standard
I	127	7793.7	20.7				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		97.401	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 14:24:18
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.250
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.340
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036211
Sample Date/Time: Tuesday, June 28, 2016 14:25:14
Number of Replicates: 3
Autosampler Position: 213
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100807.6	1.2				ug/L	81760	Standard
	Be	9	76.7	13.6	**0.0528**	0.008	15.0	ug/L	17	Standard
	Al	27	23829161.3	1.5	**125.3655**	0.841	0.7	ug/L	1410	Standard
	Sc	45	39760.5	3.7				ug/L	39543	Standard
	Ti	47	180.7	0.8	**0.5860**	0.014	2.4	ug/L	39	Standard
	V	51	2583.8	5.2	**0.1775**	0.016	9.1	ug/L	1415	Standard
	Cr	52	8684.5	0.7	**0.2325**	0.005	2.3	ug/L	6946	Standard
	Cr	53	1498.4	3.7	**1.4227**	0.048	3.4	ug/L	328	Standard
	Mn	55	44479.2	1.1	**6.9441**	0.116	1.7	ug/L	823	Standard
	Co	59	1598.1	1.9	**0.2038**	0.006	2.9	ug/L	252	Standard
	Ni	60	3161.3	2.5	**1.6303**	0.047	2.9	ug/L	192	Standard
	Cu	65	697.7	3.7	**0.2888**	0.018	6.1	ug/L	141	Standard
	Zn	66	2205.8	4.6	**1.2527**	0.096	7.7	ug/L	491	Standard
>	**Ge**	72	498311.0	1.0				ug/L	515087	Standard
	As	75	175.3	10.2	**0.2014**	0.015	7.3	ug/L	20	Standard
	Se	82	90.2	6.0	**0.6901**	0.041	6.0	ug/L	15	Standard
	Se-1	77	104.3	15.6	**0.7141**	0.236	33.0	ug/L	48	Standard
>	**Ga**	71	91.7	19.2				mg/L	90	Standard
	Rb	85	3617.1	1.0				ug/L	23	Standard
	Y	89	438807.3	1.1				ug/L	448510	Standard
>	**Rh**	103	23.3	24.7				ug/L	15	Standard
	Mo	98	5340.3	1.5	**1.9620**	0.035	1.8	ug/L	11	Standard
	Ag	107	313.7	8.5	**0.0307**	0.004	14.3	ug/L	134	Standard
	Cd	111	102.5	10.0	**0.0451**	0.004	9.9	mg/L	7	Standard
	Cd	114	269.5	9.6	**0.0471**	0.005	11.0	ug/L	16	Standard
>	**In**	115	609490.3	1.0				ug/L	611117	Standard
	Sn	118	888.4	6.3	**0.0488**	0.011	22.0	ug/L	518	Standard
	Sb	123	739.4	5.4	**0.1443**	0.009	6.4	ug/L	127	Standard
	Ba	135	23651.8	1.3	**11.4620**	0.117	1.0	ug/L	45	Standard
	Ce	140	2515.2	8.7				ug/L	298	Standard
>	**Tb**	159	994391.3	1.1				ug/L	977942	Standard
	Ho	165	61.7	16.9				ug/L	12	Standard
	Tl	203	495.7	2.3	**0.0524**	0.001	2.2	ug/L	31	Standard
	Tl	205	451.7	6.1	**0.0577**	0.003	5.9	ug/L	20	Standard
	Pb	206	782.0	4.2	**0.0770**	0.005	7.0	ug/L	286	Standard
	Pb	207	638.7	5.2	**0.0693**	0.006	8.1	ug/L	257	Standard
	Pb	208	2612.4	2.6	**0.0742**	0.003	3.5	ug/L	956	Standard
	U	238	3078.0	1.8	**0.4146**	0.006	1.5	ug/L	17	Standard
>	**Bi**	209	573149.3	0.8				ug/L	557080	Standard

Sample ID: L1606036211
Report Date/Time: Tuesday, June 28, 2016 14:27:31
Page 1

Approved: June 29, 2016

Na	23	10.0	50.0	**63.1287**	29.742	47.1	mg/L	0	Standard
Mg	24	8182.2	4.0	**10.7839**	0.556	5.2	mg/L	32	Standard
K	39	235.0	11.8	**0.6477**	0.082	12.7	mg/L	30	Standard
Ca	43	488.3	8.8	**68.4110**	5.631	8.2	mg/L	48	Standard
Fe	54	190.4	22.3	**0.0341**	0.036	104.7	mg/L	159	Standard
Fe	57	451.7	3.2	**0.8921**	0.019	2.1	mg/L	195	Standard
> **Sc-1**	45	39760.5	3.7				mg/L	39543	Standard
Cl	35	36572.1	2.1				ug/L	27642	Standard
Kr	83	2.0	0.0				ug/L	2	Standard
Br	81	15501.3	5.9				ug/L	1647	Standard
P	31	22299.7	2.1				ug/L	15231	Standard
S	34	2183.5	4.1				ug/L	1845	Standard
Sr	88	141.7	11.3				ug/L	70	Standard
C	12	286.7	23.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	128.1	22.2				mg/L	12	Standard
Ho-1	165	61.7	16.9				mg/L	12	Standard
Er	166	110.0	63.6				mg/L	20	Standard
I	127	9539.7	2.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		123.297	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.743	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036211
Report Date/Time: Tuesday, June 28, 2016 14:27:31
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	99.734
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.884
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606036211
Report Date/Time: Tuesday, June 28, 2016 14:27:31
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606036212
Sample Date/Time: Tuesday, June 28, 2016 14:28:25
Number of Replicates: 3
Autosampler Position: 214
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101556.0	2.3				ug/L	81760	Standard
	Be	9	16.7	34.6	**0.0037**	0.005	121.7	ug/L	17	Standard
	Al	27	17765744.4	2.3	**92.8094**	3.109	3.3	ug/L	1410	Standard
	Sc	45	39087.0	2.2				ug/L	39543	Standard
	Ti	47	247.0	6.3	**0.8686**	0.057	6.6	ug/L	39	Standard
	V	51	2382.1	6.3	**0.1501**	0.021	13.9	ug/L	1415	Standard
	Cr	52	8300.6	1.1	**0.1842**	0.008	4.2	ug/L	6946	Standard
	Cr	53	1501.7	3.0	**1.4394**	0.046	3.2	ug/L	328	Standard
	Mn	55	1258376.0	0.7	**201.9190**	0.817	0.4	ug/L	823	Standard
	Co	59	3356.7	1.5	**0.4660**	0.012	2.5	ug/L	252	Standard
	Ni	60	3550.4	0.9	**1.8576**	0.023	1.3	ug/L	192	Standard
	Cu	65	1089.7	4.5	**0.5061**	0.022	4.4	ug/L	141	Standard
	Zn	66	2532.2	4.9	**1.5769**	0.099	6.3	ug/L	491	Standard
>	**Ge**	72	494799.1	0.8				ug/L	515087	Standard
	As	75	371.2	4.5	**0.3840**	0.014	3.7	ug/L	20	Standard
	Se	82	96.6	14.9	**0.7542**	0.127	16.9	ug/L	15	Standard
	Se-1	77	117.3	12.8	**0.9020**	0.200	22.2	ug/L	48	Standard
>	**Ga**	71	161.7	20.6				mg/L	90	Standard
	Rb	85	5996.2	5.2				ug/L	23	Standard
	Y	89	446186.9	1.4				ug/L	448510	Standard
>	**Rh**	103	35.0	37.8				ug/L	15	Standard
	Mo	98	1982.4	1.7	**0.7175**	0.021	3.0	ug/L	11	Standard
	Ag	107	87.7	6.7	**-0.0031**	0.001	30.7	ug/L	134	Standard
	Cd	111	81.3	17.6	**0.0353**	0.007	20.5	mg/L	7	Standard
	Cd	114	230.7	12.0	**0.0397**	0.006	14.9	ug/L	16	Standard
>	**In**	115	608659.5	1.3				ug/L	611117	Standard
	Sn	118	906.7	3.5	**0.0521**	0.007	12.9	ug/L	518	Standard
	Sb	123	394.5	7.6	**0.0711**	0.006	8.6	ug/L	127	Standard
	Ba	135	51224.8	1.3	**24.8890**	0.366	1.5	ug/L	45	Standard
	Ce	140	3013.6	4.4				ug/L	298	Standard
>	**Tb**	159	974496.5	1.8				ug/L	977942	Standard
	Ho	165	80.0	16.5				ug/L	12	Standard
	Tl	203	128.3	8.3	**0.0104**	0.001	10.4	ug/L	31	Standard
	Tl	205	91.7	22.0	**0.0102**	0.003	28.0	ug/L	20	Standard
	Pb	206	755.4	5.1	**0.0757**	0.008	10.8	ug/L	286	Standard
	Pb	207	644.0	3.0	**0.0736**	0.005	7.3	ug/L	257	Standard
	Pb	208	2468.4	2.8	**0.0700**	0.005	7.5	ug/L	956	Standard
	U	238	1502.1	3.5	**0.2059**	0.006	2.8	ug/L	17	Standard
>	**Bi**	209	559159.1	1.2				ug/L	557080	Standard

Sample ID: L1606036212
Report Date/Time: Tuesday, June 28, 2016 14:30:42
Page 1

Approved: June 29, 2016

Na	23	5.0	0.0	**32.4897**	0.713	2.2	mg/L	0	Standard
Mg	24	34367.3	2.9	**46.2225**	0.714	1.5	mg/L	32	Standard
K	39	518.3	10.6	**1.6321**	0.206	12.7	mg/L	30	Standard
Ca	43	333.3	3.5	**45.8704**	1.237	2.7	mg/L	48	Standard
Fe	54	313.8	12.6	**0.1508**	0.042	28.0	mg/L	159	Standard
Fe	57	493.3	3.3	**1.0616**	0.089	8.4	mg/L	195	Standard
> **Sc-1**	45	39087.0	2.2				mg/L	39543	Standard
Cl	35	40128.4	3.1				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	13105.7	7.2				ug/L	1647	Standard
P	31	21068.0	1.2				ug/L	15231	Standard
S	34	2031.8	3.8				ug/L	1845	Standard
Sr	88	195.0	14.3				ug/L	70	Standard
C	12	220.0	19.8				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	113.5	27.0				mg/L	12	Standard
Ho-1	165	80.0	16.5				mg/L	12	Standard
Er	166	66.7	48.2				mg/L	20	Standard
I	127	14735.5	4.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		124.212	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.061	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606036212
Report Date/Time: Tuesday, June 28, 2016 14:30:42
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.598
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.373
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 14:31:38
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	96993.2	4.5				ug/L	81760	Standard
	Be	9	59018.4	6.5	**50.0965**	2.769	5.5	ug/L	17	Standard
	Al	27	7977258.4	4.2	**43.6214**	1.029	2.4	ug/L	1410	Standard
	Sc	45	38238.2	1.6				ug/L	39543	Standard
	Ti	47	23612.4	4.0	**98.7594**	3.955	4.0	ug/L	39	Standard
	V	51	340072.0	1.7	**50.7725**	1.247	2.5	ug/L	1415	Standard
	Cr	52	338200.8	3.5	**49.5590**	1.660	3.3	ug/L	6946	Standard
	Cr	53	42016.5	3.2	**49.5286**	1.703	3.4	ug/L	328	Standard
	Mn	55	307360.3	1.5	**49.2454**	0.735	1.5	ug/L	823	Standard
	Co	59	358545.3	2.7	**53.1356**	1.497	2.8	ug/L	252	Standard
	Ni	60	89705.8	2.1	**49.5206**	1.132	2.3	ug/L	192	Standard
	Cu	65	91128.3	2.1	**49.8745**	1.003	2.0	ug/L	141	Standard
	Zn	66	53559.8	1.4	**50.0843**	0.931	1.9	ug/L	491	Standard
>	Ge	72	494423.7	1.0				ug/L	515087	Standard
	As	75	53994.0	0.9	**50.0733**	0.496	1.0	ug/L	20	Standard
	Se	82	5715.9	0.2	**51.6065**	0.476	0.9	ug/L	15	Standard
	Se-1	77	3736.8	1.4	**50.9602**	0.952	1.9	ug/L	48	Standard
>	Ga	71	70.0	28.6				mg/L	90	Standard
	Rb	85	770.0	3.6				ug/L	23	Standard
	Y	89	445528.9	0.6				ug/L	448510	Standard
>	Rh	103	23.3	44.6				ug/L	15	Standard
	Mo	98	264652.3	1.1	**96.1672**	0.791	0.8	ug/L	11	Standard
	Ag	107	331013.3	2.8	**48.4244**	1.155	2.4	ug/L	134	Standard
	Cd	111	104542.0	2.3	**48.0434**	1.120	2.3	mg/L	7	Standard
	Cd	114	255433.5	4.1	**48.2912**	1.921	4.0	ug/L	16	Standard
>	In	115	622075.2	0.5				ug/L	611117	Standard
	Sn	118	295472.2	1.9	**48.9864**	0.743	1.5	ug/L	518	Standard
	Sb	123	230757.9	1.7	**48.0319**	0.618	1.3	ug/L	127	Standard
	Ba	135	99769.8	2.2	**47.4498**	0.988	2.1	ug/L	45	Standard
	Ce	140	73.3	7.9				ug/L	298	Standard
>	Tb	159	1000112.3	1.5				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	427237.6	2.4	**49.0359**	1.247	2.5	ug/L	31	Standard
	Tl	205	388094.9	2.7	**51.3011**	1.717	3.3	ug/L	20	Standard
	Pb	206	274003.7	2.0	**48.9721**	1.261	2.6	ug/L	286	Standard
	Pb	207	249924.4	1.5	**49.5263**	1.133	2.3	ug/L	257	Standard
	Pb	208	988217.1	0.7	**50.6156**	0.799	1.6	ug/L	956	Standard
	U	238	380024.3	2.2	**51.3005**	1.621	3.2	ug/L	17	Standard
>	Bi	209	576058.0	1.3				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:33:55
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**33.2699**	32.995	99.2	mg/L	0	Standard
Mg	24	4203.9	5.4	**5.7315**	0.340	5.9	mg/L	32	Standard
K	39	1220.0	6.7	**4.1203**	0.250	6.1	mg/L	30	Standard
Ca	43	103.3	14.8	**10.9126**	2.675	24.5	mg/L	48	Standard
Fe	54	5915.4	3.1	**5.3996**	0.173	3.2	mg/L	159	Standard
Fe	57	1661.8	2.4	**5.1766**	0.085	1.6	mg/L	195	Standard
> **Sc-1**	45	38238.2	1.6				mg/L	39543	Standard
Cl	35	33728.9	2.6				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	1893.5	8.5				ug/L	1647	Standard
P	31	18936.8	1.0				ug/L	15231	Standard
S	34	2451.9	4.7				ug/L	1845	Standard
Sr	88	73.3	14.2				ug/L	70	Standard
C	12	160.0	43.3				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.2	6.0				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	16.7	69.3				mg/L	20	Standard
I	127	3007.0	16.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	100.193		
Al	27	87.243		
Sc	45			
Ti	47	98.759		
V	51	101.545		
Cr	52	99.118		
Cr	53			
Mn	55	98.491		
Co	59	106.271		
Ni	60	99.041		
Cu	65	99.749		
Zn	66	100.169		
> Ge	72		95.988	
As	75	100.147		
Se	82	103.213		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:33:55
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	96.167		
Ag	107	96.849		
Cd	111	96.087		
Cd	114			
In	115			101.793
Sn	118	97.973		
Sb	123	96.064		
Ba	135	94.900		
Ce	140			
Tb	159			
Ho	165			
Tl	203	98.072		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.231		
U	238	102.601		
Bi	209			103.407
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 14:33:55
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 14:34:50
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92870.5	2.0				ug/L	81760	Standard
	Be	9	15.0	66.7	**0.0036**	0.009	245.7	ug/L	17	Standard
	Al	27	2743.6	21.7	**0.0013**	0.004	284.3	ug/L	1410	Standard
	Sc	45	38460.4	1.3				ug/L	39543	Standard
	Ti	47	30.7	8.2	**-0.0403**	0.010	23.8	ug/L	39	Standard
	V	51	1134.9	2.6	**-0.0414**	0.005	11.0	ug/L	1415	Standard
	Cr	52	5878.1	2.2	**-0.2023**	0.026	12.8	ug/L	6946	Standard
	Cr	53	310.0	1.6	**0.0160**	0.005	29.1	ug/L	328	Standard
	Mn	55	709.4	9.1	**-0.0377**	0.010	26.9	ug/L	823	Standard
	Co	59	233.7	1.1	**0.0023**	0.001	24.5	ug/L	252	Standard
	Ni	60	211.0	9.6	**0.0084**	0.011	126.5	ug/L	192	Standard
	Cu	65	155.3	8.6	**-0.0081**	0.008	96.4	ug/L	141	Standard
	Zn	66	394.0	4.1	**-0.4638**	0.014	3.1	ug/L	491	Standard
>	**Ge**	72	508932.6	0.9				ug/L	515087	Standard
	As	75	-34.1	13.4	**0.0096**	0.004	45.1	ug/L	20	Standard
	Se	82	19.4	20.9	**0.0515**	0.036	69.8	ug/L	15	Standard
	Se-1	77	46.7	12.2	**-0.0918**	0.077	84.0	ug/L	48	Standard
>	**Ga**	71	66.7	8.7				mg/L	90	Standard
	Rb	85	15.0	33.3				ug/L	23	Standard
	Y	89	445570.9	1.5				ug/L	448510	Standard
>	**Rh**	103	15.0	88.2				ug/L	15	Standard
	Mo	98	287.3	16.9	**0.0864**	0.017	20.0	ug/L	11	Standard
	Ag	107	124.0	22.6	**0.0021**	0.004	200.7	ug/L	134	Standard
	Cd	111	14.8	9.5	**0.0040**	0.001	16.6	mg/L	7	Standard
	Cd	114	48.0	37.7	**0.0043**	0.004	82.8	ug/L	16	Standard
>	**In**	115	615541.5	0.9				ug/L	611117	Standard
	Sn	118	883.4	22.9	**0.0465**	0.035	74.9	ug/L	518	Standard
	Sb	123	479.5	25.7	**0.0882**	0.027	30.1	ug/L	127	Standard
	Ba	135	35.7	10.6	**-0.0080**	0.002	22.2	ug/L	45	Standard
	Ce	140	21.7	48.0				ug/L	298	Standard
>	**Tb**	159	980265.4	0.4				ug/L	977942	Standard
	Ho	165	16.7	45.8				ug/L	12	Standard
	Tl	203	63.0	23.1	**0.0024**	0.002	67.2	ug/L	31	Standard
	Tl	205	63.3	31.9	**0.0060**	0.003	42.7	ug/L	20	Standard
	Pb	206	313.3	5.4	**-0.0080**	0.003	36.6	ug/L	286	Standard
	Pb	207	273.0	6.7	**-0.0043**	0.004	87.0	ug/L	257	Standard
	Pb	208	1087.7	7.3	**-0.0051**	0.004	85.3	ug/L	956	Standard
	U	238	49.0	30.5	**0.0036**	0.002	56.1	ug/L	17	Standard
>	**Bi**	209	580619.2	1.2				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 14:37:07
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	35.0	28.6	**-0.0091**	0.014	156.5	mg/L	32	Standard
K	39	36.7	28.4	**-0.0142**	0.035	248.1	mg/L	30	Standard
Ca	43	26.7	57.3	**-1.1996**	2.339	195.0	mg/L	48	Standard
Fe	54	140.8	15.9	**-0.0061**	0.021	341.9	mg/L	159	Standard
Fe	57	151.7	6.9	**-0.0978**	0.042	43.3	mg/L	195	Standard
> **Sc-1**	45	38460.4	1.3				mg/L	39543	Standard
Cl	35	32500.2	3.2				ug/L	27642	Standard
Kr	83	3.0	0.0				ug/L	2	Standard
Br	81	1736.8	3.3				ug/L	1647	Standard
P	31	17698.7	3.2				ug/L	15231	Standard
S	34	2320.2	10.5				ug/L	1845	Standard
Sr	88	73.3	21.9				ug/L	70	Standard
C	12	133.3	15.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	9.0	113.3				mg/L	12	Standard
Ho-1	165	16.7	45.8				mg/L	12	Standard
Er	166	20.0	50.0				mg/L	20	Standard
I	127	2206.8	20.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.805	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 14:37:07
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.724
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.225
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 14:37:07
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: PBS M7 WG573333-03
Sample Date/Time: Tuesday, June 28, 2016 14:38:03
Number of Replicates: 3
Autosampler Position: 215
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95050.7	3.1				ug/L	81760	Standard
	Be	9	18.3	87.7	**0.0062**	0.014	223.6	ug/L	17	Standard
	Al	27	17488.4	2.4	**0.0832**	0.003	3.2	ug/L	1410	Standard
	Sc	45	39132.2	2.7				ug/L	39543	Standard
	Ti	47	49.7	51.3	**0.0353**	0.101	286.1	ug/L	39	Standard
	V	51	1251.1	3.7	**-0.0255**	0.008	31.0	ug/L	1415	Standard
	Cr	52	8706.2	0.9	**0.2013**	0.013	6.2	ug/L	6946	Standard
	Cr	53	700.0	11.0	**0.4613**	0.093	20.2	ug/L	328	Standard
	Mn	55	1542.7	1.5	**0.0910**	0.006	6.6	ug/L	823	Standard
	Co	59	278.7	9.1	**0.0085**	0.003	38.3	ug/L	252	Standard
	Ni	60	602.0	5.9	**0.2166**	0.018	8.1	ug/L	192	Standard
	Cu	65	210.7	10.6	**0.0207**	0.012	58.0	ug/L	141	Standard
	Zn	66	1817.8	2.7	**0.8408**	0.028	3.4	ug/L	491	Standard
>	**Ge**	72	511962.6	1.0				ug/L	515087	Standard
	As	75	-35.9	60.1	**0.0082**	0.019	234.7	ug/L	20	Standard
	Se	82	20.1	18.9	**0.0563**	0.034	59.9	ug/L	15	Standard
	Se-1	77	44.7	10.1	**-0.1225**	0.058	47.0	ug/L	48	Standard
>	**Ga**	71	48.3	15.8				mg/L	90	Standard
	Rb	85	66.7	31.2				ug/L	23	Standard
	Y	89	442200.8	1.3				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	193.0	16.8	**0.0509**	0.011	21.8	ug/L	11	Standard
	Ag	107	97.0	17.0	**-0.0021**	0.002	110.1	ug/L	134	Standard
	Cd	111	7.4	64.2	**0.0005**	0.002	466.2	mg/L	7	Standard
	Cd	114	31.4	28.6	**0.0010**	0.002	163.8	ug/L	16	Standard
>	**In**	115	622890.2	1.7				ug/L	611117	Standard
	Sn	118	691.7	5.6	**0.0128**	0.005	38.1	ug/L	518	Standard
	Sb	123	212.1	15.5	**0.0312**	0.006	20.3	ug/L	127	Standard
	Ba	135	298.3	2.4	**0.1167**	0.001	1.2	ug/L	45	Standard
	Ce	140	113.3	24.3				ug/L	298	Standard
>	**Tb**	159	998579.7	1.5				ug/L	977942	Standard
	Ho	165	16.7	17.3				ug/L	12	Standard
	Tl	203	39.0	48.9	**-0.0004**	0.002	526.4	ug/L	31	Standard
	Tl	205	20.0	50.0	**0.0003**	0.001	421.9	ug/L	20	Standard
	Pb	206	313.7	9.3	**-0.0086**	0.005	56.5	ug/L	286	Standard
	Pb	207	273.0	2.4	**-0.0050**	0.001	28.4	ug/L	257	Standard
	Pb	208	1093.3	1.9	**-0.0055**	0.001	16.3	ug/L	956	Standard
	U	238	24.3	55.2	**0.0002**	0.002	711.6	ug/L	17	Standard
>	**Bi**	209	587793.8	0.5				ug/L	557080	Standard

Sample ID: PBS M7 WG573333-03
Report Date/Time: Tuesday, June 28, 2016 14:40:20
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	36.7	7.9	**-0.0078**	0.003	35.7	mg/L	32	Standard
K	39	35.0	51.5	**-0.0208**	0.064	309.0	mg/L	30	Standard
Ca	43	25.0	34.6	**-1.5082**	1.327	88.0	mg/L	48	Standard
Fe	54	165.5	15.5	**0.0140**	0.019	138.8	mg/L	159	Standard
Fe	57	203.3	9.9	**0.0704**	0.084	119.9	mg/L	195	Standard
> **Sc-1**	45	39132.2	2.7				mg/L	39543	Standard
Cl	35	33734.8	0.7				ug/L	27642	Standard
Kr	83	2.3	65.5				ug/L	2	Standard
Br	81	1786.8	3.4				ug/L	1647	Standard
P	31	17722.0	1.9				ug/L	15231	Standard
S	34	2341.8	1.1				ug/L	1845	Standard
Sr	88	65.0	27.7				ug/L	70	Standard
C	12	180.0	47.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	18.7	93.3				mg/L	12	Standard
Ho-1	165	16.7	17.3				mg/L	12	Standard
Er	166	26.7	21.7				mg/L	20	Standard
I	127	1696.8	8.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		116.256	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.393	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBS M7 WG573333-03
Report Date/Time: Tuesday, June 28, 2016 14:40:20
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.927
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.513
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: LCSS M7 WG573333-04
Sample Date/Time: Tuesday, June 28, 2016 14:41:14
Number of Replicates: 3
Autosampler Position: 216
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92927.6	3.1				ug/L	81760	Standard
	Be	9	30111.6	3.7	**26.7031**	1.769	6.6	ug/L	17	Standard
	Al	27	18035.9	13.5	**0.0888**	0.016	18.2	ug/L	1410	Standard
	Sc	45	39361.1	1.9				ug/L	39543	Standard
	Ti	47	37.0	28.2	**-0.0130**	0.042	322.2	ug/L	39	Standard
	V	51	182372.3	3.8	**26.6531**	1.518	5.7	ug/L	1415	Standard
	Cr	52	189442.3	3.6	**26.8012**	1.501	5.6	ug/L	6946	Standard
	Cr	53	22922.3	1.8	**26.3866**	1.122	4.3	ug/L	328	Standard
	Mn	55	169435.1	3.2	**26.6017**	1.285	4.8	ug/L	823	Standard
	Co	59	193981.0	3.8	**28.2331**	1.651	5.8	ug/L	252	Standard
	Ni	60	48862.7	3.3	**26.4525**	1.352	5.1	ug/L	192	Standard
	Cu	65	49743.6	2.7	**26.7036**	1.173	4.4	ug/L	141	Standard
	Zn	66	30362.1	3.5	**27.5261**	1.455	5.3	ug/L	491	Standard
>	**Ge**	72	503563.1	2.3				ug/L	515087	Standard
	As	75	29155.1	3.3	**26.5827**	1.325	5.0	ug/L	20	Standard
	Se	82	3055.7	1.7	**27.0468**	1.107	4.1	ug/L	15	Standard
	Se-1	77	2068.5	3.5	**27.3757**	1.137	4.2	ug/L	48	Standard
>	**Ga**	71	53.3	14.3				mg/L	90	Standard
	Rb	85	78.3	16.1				ug/L	23	Standard
	Y	89	444634.7	2.0				ug/L	448510	Standard
>	**Rh**	103	18.3	41.7				ug/L	15	Standard
	Mo	98	104.5	23.9	**0.0190**	0.010	51.3	ug/L	11	Standard
	Ag	107	179344.1	3.7	**26.2092**	1.706	6.5	ug/L	134	Standard
	Cd	111	55845.5	2.3	**25.6355**	1.334	5.2	mg/L	7	Standard
	Cd	114	136877.4	2.6	**25.8483**	1.371	5.3	ug/L	16	Standard
>	**In**	115	623373.0	3.0				ug/L	611117	Standard
	Sn	118	755.0	11.3	**0.0235**	0.016	68.2	ug/L	518	Standard
	Sb	123	124890.1	3.5	**25.9676**	1.618	6.2	ug/L	127	Standard
	Ba	135	53944.6	3.3	**25.6212**	1.575	6.1	ug/L	45	Standard
	Ce	140	106.7	9.8				ug/L	298	Standard
>	**Tb**	159	976375.9	3.3				ug/L	977942	Standard
	Ho	165	15.0	57.7				ug/L	12	Standard
	Tl	203	229696.9	2.8	**26.5184**	1.549	5.8	ug/L	31	Standard
	Tl	205	205993.8	3.8	**27.3937**	1.852	6.8	ug/L	20	Standard
	Pb	206	150555.8	3.2	**27.0395**	1.637	6.1	ug/L	286	Standard
	Pb	207	130667.9	3.7	**26.0217**	1.699	6.5	ug/L	257	Standard
	Pb	208	532429.0	3.5	**27.4071**	1.756	6.4	ug/L	956	Standard
	U	238	197217.2	3.4	**26.7798**	1.720	6.4	ug/L	17	Standard
>	**Bi**	209	573273.6	3.1				ug/L	557080	Standard

Sample ID: LCSS M7 WG573333-04
Report Date/Time: Tuesday, June 28, 2016 14:43:31
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**21.2655**	18.412	86.6	mg/L	0	Standard
Mg	24	50.0	43.6	**0.0096**	0.028	296.2	mg/L	32	Standard
K	39	30.0	100.0	**-0.0395**	0.101	254.9	mg/L	30	Standard
Ca	43	25.0	40.0	**-1.5307**	1.509	98.6	mg/L	48	Standard
Fe	54	141.7	2.7	**-0.0082**	0.003	33.7	mg/L	159	Standard
Fe	57	226.7	9.9	**0.1435**	0.065	45.0	mg/L	195	Standard
> **Sc-1**	45	39361.1	1.9				mg/L	39543	Standard
Cl	35	35061.2	2.7				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1886.8	4.9				ug/L	1647	Standard
P	31	18956.9	6.3				ug/L	15231	Standard
S	34	2375.2	1.6				ug/L	1845	Standard
Sr	88	65.0	30.8				ug/L	70	Standard
C	12	170.0	38.6				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	12.7	49.8				mg/L	12	Standard
Ho-1	165	15.0	57.7				mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	1553.4	15.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.659	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.763	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSS M7 WG573333-04
Report Date/Time: Tuesday, June 28, 2016 14:43:31
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.006
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.907
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSS M7 WG573333-04
Report Date/Time: Tuesday, June 28, 2016 14:43:31
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606105301 WG573333-02
Sample Date/Time: Tuesday, June 28, 2016 14:44:26
Number of Replicates: 3
Autosampler Position: 217
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92504.9	1.8				ug/L	81760	Standard
	Be	9	211.7	3.6	**0.1787**	0.009	5.2	ug/L	17	Standard
	Al	27	476632.7	0.6	**2.7192**	0.058	2.1	ug/L	1410	Standard
	Sc	45	39899.1	1.7				ug/L	39543	Standard
	Ti	47	7508.9	2.4	**29.9625**	0.867	2.9	ug/L	39	Standard
	V	51	54226.8	0.3	**7.5775**	0.105	1.4	ug/L	1415	Standard
	Cr	52	35510.9	0.6	**4.0327**	0.059	1.5	ug/L	6946	Standard
	Cr	53	3988.9	7.5	**4.1910**	0.323	7.7	ug/L	328	Standard
	Mn	55	226862.1	2.5	**34.7724**	1.274	3.7	ug/L	823	Standard
	Co	59	8803.9	0.2	**1.2188**	0.017	1.4	ug/L	252	Standard
	Ni	60	5989.2	1.6	**3.0677**	0.039	1.3	ug/L	192	Standard
	Cu	65	4442.7	0.7	**2.2418**	0.014	0.6	ug/L	141	Standard
	Zn	66	19698.8	1.0	**17.1048**	0.367	2.1	ug/L	491	Standard
>	**Ge**	72	516312.9	1.2				ug/L	515087	Standard
	As	75	2960.6	0.8	**2.6677**	0.048	1.8	ug/L	20	Standard
	Se	82	53.9	7.9	**0.3477**	0.042	12.1	ug/L	15	Standard
	Se-1	77	82.0	6.8	**0.3676**	0.085	23.1	ug/L	48	Standard
>	**Ga**	71	5075.9	3.2				mg/L	90	Standard
	Rb	85	19981.5	3.8				ug/L	23	Standard
	Y	89	472017.4	2.4				ug/L	448510	Standard
>	**Rh**	103	21.7	26.6				ug/L	15	Standard
	Mo	98	6655.2	1.4	**2.4116**	0.043	1.8	ug/L	11	Standard
	Ag	107	178.7	11.3	**0.0101**	0.003	29.0	ug/L	134	Standard
	Cd	111	244.9	9.1	**0.1102**	0.010	9.1	mg/L	7	Standard
	Cd	114	621.2	5.8	**0.1131**	0.007	6.2	ug/L	16	Standard
>	**In**	115	619055.2	0.3				ug/L	611117	Standard
	Sn	118	640.0	5.5	**0.0049**	0.006	124.7	ug/L	518	Standard
	Sb	123	463.5	5.3	**0.0841**	0.005	6.4	ug/L	127	Standard
	Ba	135	41539.5	1.0	**19.8378**	0.203	1.0	ug/L	45	Standard
	Ce	140	324520.8	1.6				ug/L	298	Standard
>	**Tb**	159	988287.6	1.4				ug/L	977942	Standard
	Ho	165	2263.5	6.3				ug/L	12	Standard
	Tl	203	550.3	1.1	**0.0581**	0.001	1.2	ug/L	31	Standard
	Tl	205	506.7	1.1	**0.0644**	0.001	1.1	ug/L	20	Standard
	Pb	206	143402.5	0.9	**25.4978**	0.319	1.3	ug/L	286	Standard
	Pb	207	120708.6	0.6	**23.7943**	0.166	0.7	ug/L	257	Standard
	Pb	208	491539.2	1.4	**25.0461**	0.389	1.6	ug/L	956	Standard
	U	238	1919.1	2.3	**0.2551**	0.006	2.5	ug/L	17	Standard
>	**Bi**	209	578289.7	0.5				ug/L	557080	Standard

Sample ID: L1606105301 WG573333-02
Report Date/Time: Tuesday, June 28, 2016 14:46:43
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.6918**	18.510	173.1	mg/L	0	Standard
Mg	24	41.7	30.2	**-0.0023**	0.016	678.2	mg/L	32	Standard
K	39	48.3	15.8	**0.0204**	0.026	126.1	mg/L	30	Standard
Ca	43	38.3	27.2	**0.4078**	1.458	357.4	mg/L	48	Standard
Fe	54	3682.2	3.4	**3.1665**	0.134	4.2	mg/L	159	Standard
Fe	57	1100.0	9.6	**3.0601**	0.408	13.3	mg/L	195	Standard
> **Sc-1**	45	39899.1	1.7				mg/L	39543	Standard
Cl	35	32895.7	2.7				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	2153.5	9.2				ug/L	1647	Standard
P	31	17883.9	3.0				ug/L	15231	Standard
S	34	2195.2	9.1				ug/L	1845	Standard
Sr	88	71.7	4.0				ug/L	70	Standard
C	12	220.0	19.8				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	40.0	25.0				mg/L	0	Standard
Dy	164	4107.1	8.5				mg/L	12	Standard
Ho-1	165	2263.5	6.3				mg/L	12	Standard
Er	166	2160.2	2.1				mg/L	20	Standard
I	127	7528.5	2.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.142	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.238	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606105301 WG573333-02
Report Date/Time: Tuesday, June 28, 2016 14:46:43
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.299
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.807
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606105302S WG573333-07
Sample Date/Time: Tuesday, June 28, 2016 14:47:37
Number of Replicates: 3
Autosampler Position: 218
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	91483.8	2.5				ug/L	81760	Standard
	Be	9	28575.3	3.7	**25.7021**	0.502	2.0	ug/L	17	Standard
	Al	27	472647.5	0.9	**2.7267**	0.044	1.6	ug/L	1410	Standard
	Sc	45	40355.4	2.6				ug/L	39543	Standard
	Ti	47	8788.9	2.8	**34.8074**	0.551	1.6	ug/L	39	Standard
	V	51	216754.8	1.8	**30.6603**	0.748	2.4	ug/L	1415	Standard
	Cr	52	202224.9	0.6	**27.6955**	0.221	0.8	ug/L	6946	Standard
	Cr	53	25461.4	2.5	**28.3619**	0.467	1.6	ug/L	328	Standard
	Mn	55	374661.4	3.0	**57.0462**	1.611	2.8	ug/L	823	Standard
	Co	59	185021.8	2.1	**26.0317**	0.592	2.3	ug/L	252	Standard
	Ni	60	51684.0	1.6	**27.0535**	0.107	0.4	ug/L	192	Standard
	Cu	65	52133.2	1.4	**27.0623**	0.171	0.6	ug/L	141	Standard
	Zn	66	46187.6	0.8	**40.8778**	0.267	0.7	ug/L	491	Standard
>	**Ge**	72	520472.6	1.2				ug/L	515087	Standard
	As	75	27327.3	2.3	**24.0946**	0.465	1.9	ug/L	20	Standard
	Se	82	2628.8	2.4	**22.4788**	0.502	2.2	ug/L	15	Standard
	Se-1	77	1784.1	1.0	**22.7201**	0.347	1.5	ug/L	48	Standard
>	**Ga**	71	5215.9	5.7				mg/L	90	Standard
	Rb	85	16937.8	2.3				ug/L	23	Standard
	Y	89	458647.4	1.3				ug/L	448510	Standard
>	**Rh**	103	30.0	44.1				ug/L	15	Standard
	Mo	98	18905.8	1.4	**6.8731**	0.106	1.5	ug/L	11	Standard
	Ag	107	155307.6	0.6	**22.7819**	0.123	0.5	ug/L	134	Standard
	Cd	111	52492.7	0.7	**24.1959**	0.317	1.3	mg/L	7	Standard
	Cd	114	128079.2	0.4	**24.2856**	0.076	0.3	ug/L	16	Standard
>	**In**	115	620204.7	0.6				ug/L	611117	Standard
	Sn	118	685.0	20.9	**0.0121**	0.023	191.0	ug/L	518	Standard
	Sb	123	65087.0	0.8	**13.5800**	0.099	0.7	ug/L	127	Standard
	Ba	135	93846.7	0.6	**44.7672**	0.375	0.8	ug/L	45	Standard
	Ce	140	331961.2	1.7				ug/L	298	Standard
>	**Tb**	159	981697.8	0.7				ug/L	977942	Standard
	Ho	165	1903.5	2.0				ug/L	12	Standard
	Tl	203	213525.6	1.3	**24.5875**	0.365	1.5	ug/L	31	Standard
	Tl	205	191177.1	1.7	**25.3529**	0.481	1.9	ug/L	20	Standard
	Pb	206	446032.3	1.2	**80.0233**	1.190	1.5	ug/L	286	Standard
	Pb	207	376762.1	1.5	**74.9381**	1.277	1.7	ug/L	257	Standard
	Pb	208	1506170.9	0.7	**77.4336**	0.690	0.9	ug/L	956	Standard
	U	238	183525.9	0.8	**24.8532**	0.261	1.1	ug/L	17	Standard
>	**Bi**	209	574082.0	0.2				ug/L	557080	Standard

Sample ID: L1606105302S WG573333-07
Report Date/Time: Tuesday, June 28, 2016 14:49:54
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	51.7	22.3	**0.0106**	0.017	161.8	mg/L	32	Standard
K	39	33.3	8.7	**-0.0312**	0.007	21.8	mg/L	30	Standard
Ca	43	38.3	15.1	**0.3768**	1.024	271.7	mg/L	48	Standard
Fe	54	2984.3	5.7	**2.5132**	0.224	8.9	mg/L	159	Standard
Fe	57	1001.7	4.5	**2.6914**	0.198	7.3	mg/L	195	Standard
> **Sc-1**	45	40355.4	2.6				mg/L	39543	Standard
Cl	35	31360.4	1.4				ug/L	27642	Standard
Kr	83	1.0	173.2				ug/L	2	Standard
Br	81	2753.6	4.6				ug/L	1647	Standard
P	31	16745.9	1.5				ug/L	15231	Standard
S	34	2118.5	5.5				ug/L	1845	Standard
Sr	88	78.3	9.8				ug/L	70	Standard
C	12	286.7	29.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	56.7	100.3				mg/L	0	Standard
Dy	164	3434.6	3.8				mg/L	12	Standard
Ho-1	165	1903.5	2.0				mg/L	12	Standard
Er	166	1860.1	3.3				mg/L	20	Standard
I	127	5512.7	1.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.893	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.046	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606105302S WG573333-07
Report Date/Time: Tuesday, June 28, 2016 14:49:54
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.487
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.052
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606105303SD WG573333-08
Sample Date/Time: Tuesday, June 28, 2016 14:50:49
Number of Replicates: 3
Autosampler Position: 219
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90707.5	2.1				ug/L	81760	Standard
	Be	9	28490.1	1.5	**25.8507**	0.239	0.9	ug/L	17	Standard
	Al	27	517242.7	0.8	**3.0108**	0.045	1.5	ug/L	1410	Standard
	Sc	45	41104.0	1.1				ug/L	39543	Standard
	Ti	47	9706.8	1.9	**39.3684**	1.018	2.6	ug/L	39	Standard
	V	51	220560.4	0.1	**31.9323**	0.545	1.7	ug/L	1415	Standard
	Cr	52	204805.0	1.1	**28.7389**	0.342	1.2	ug/L	6946	Standard
	Cr	53	25147.5	3.9	**28.6674**	1.000	3.5	ug/L	328	Standard
	Mn	55	371807.0	1.2	**57.9433**	1.578	2.7	ug/L	823	Standard
	Co	59	187737.3	0.6	**27.0316**	0.558	2.1	ug/L	252	Standard
	Ni	60	52302.5	0.6	**28.0202**	0.336	1.2	ug/L	192	Standard
	Cu	65	51530.8	0.8	**27.3738**	0.234	0.9	ug/L	141	Standard
	Zn	66	58955.2	1.6	**53.6399**	0.321	0.6	ug/L	491	Standard
>	**Ge**	72	508670.0	1.6				ug/L	515087	Standard
	As	75	26613.2	0.3	**24.0128**	0.310	1.3	ug/L	20	Standard
	Se	82	2643.5	1.0	**23.1344**	0.377	1.6	ug/L	15	Standard
	Se-1	77	1782.1	1.4	**23.2396**	0.549	2.4	ug/L	48	Standard
>	**Ga**	71	5607.7	2.4				mg/L	90	Standard
	Rb	85	19120.4	0.8				ug/L	23	Standard
	Y	89	464944.5	0.8				ug/L	448510	Standard
>	**Rh**	103	20.0	66.1				ug/L	15	Standard
	Mo	98	6589.1	2.9	**2.3569**	0.065	2.7	ug/L	11	Standard
	Ag	107	161248.5	0.9	**23.3982**	0.239	1.0	ug/L	134	Standard
	Cd	111	52808.4	0.9	**24.0772**	0.233	1.0	mg/L	7	Standard
	Cd	114	128062.2	0.6	**24.0198**	0.192	0.8	ug/L	16	Standard
>	**In**	115	626980.9	0.2				ug/L	611117	Standard
	Sn	118	691.7	6.1	**0.0121**	0.007	57.7	ug/L	518	Standard
	Sb	123	64415.1	1.9	**13.2945**	0.276	2.1	ug/L	127	Standard
	Ba	135	100181.1	0.9	**47.2728**	0.453	1.0	ug/L	45	Standard
	Ce	140	306983.9	0.9				ug/L	298	Standard
>	**Tb**	159	992773.6	1.4				ug/L	977942	Standard
	Ho	165	1925.1	2.4				ug/L	12	Standard
	Tl	203	213931.4	0.2	**24.2345**	0.027	0.1	ug/L	31	Standard
	Tl	205	188875.1	1.2	**24.6411**	0.299	1.2	ug/L	20	Standard
	Pb	206	399455.5	0.3	**70.4964**	0.248	0.4	ug/L	286	Standard
	Pb	207	335397.3	0.3	**65.6206**	0.221	0.3	ug/L	257	Standard
	Pb	208	1377394.4	0.8	**69.6587**	0.536	0.8	ug/L	956	Standard
	U	238	183479.9	0.5	**24.4438**	0.133	0.5	ug/L	17	Standard
>	**Bi**	209	583538.4	0.1				ug/L	557080	Standard

Sample ID: L1606105303SD WG573333-08
Report Date/Time: Tuesday, June 28, 2016 14:53:06
Page 1

Approved: June 29, 2016

	Na	23	0.0		**0.0050**			mg/L	0	Standard
	Mg	24	45.0	11.1	**0.0005**	0.006	1275.5	mg/L	32	Standard
	K	39	46.7	12.4	**0.0102**	0.018	179.3	mg/L	30	Standard
	Ca	43	46.7	40.6	**1.4697**	2.764	188.0	mg/L	48	Standard
	Fe	54	3552.4	2.1	**2.9565**	0.098	3.3	mg/L	159	Standard
	Fe	57	1123.4	10.8	**3.0231**	0.384	12.7	mg/L	195	Standard
>	**Sc-1**	45	41104.0	1.1				mg/L	39543	Standard
	Cl	35	31493.4	1.6				ug/L	27642	Standard
	Kr	83	1.7	69.3				ug/L	2	Standard
	Br	81	2126.8	11.6				ug/L	1647	Standard
	P	31	16392.2	2.4				ug/L	15231	Standard
	S	34	1988.5	3.7				ug/L	1845	Standard
	Sr	88	56.7	20.4				ug/L	70	Standard
	C	12	256.7	29.2				mg/L	113	Standard
	N	14	0.0					mg/L	0	Standard
	Hg	202	73.3	34.3				mg/L	0	Standard
	Dy	164	3598.5	4.6				mg/L	12	Standard
	Ho-1	165	1925.1	2.4				mg/L	12	Standard
	Er	166	1710.1	5.6				mg/L	20	Standard
	I	127	5772.8	6.4				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		110.943	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		98.754	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606105303SD WG573333-08
Report Date/Time: Tuesday, June 28, 2016 14:53:06
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.596
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.749
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606104801
Sample Date/Time: Tuesday, June 28, 2016 14:54:00
Number of Replicates: 3
Autosampler Position: 220
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87619.5	3.9				ug/L	81760	Standard
	Be	9	323.3	11.2	**0.2951**	0.044	15.0	ug/L	17	Standard
	Al	27	1546436.6	2.8	**9.3502**	0.142	1.5	ug/L	1410	Standard
	Sc	45	40666.2	2.1				ug/L	39543	Standard
	Ti	47	348.0	4.5	**1.2425**	0.055	4.4	ug/L	39	Standard
	V	51	27678.8	0.4	**3.7995**	0.033	0.9	ug/L	1415	Standard
	Cr	52	26992.7	1.5	**2.8437**	0.051	1.8	ug/L	6946	Standard
	Cr	53	2932.0	5.7	**3.0177**	0.170	5.6	ug/L	328	Standard
	Mn	55	139361.0	0.9	**21.4761**	0.101	0.5	ug/L	823	Standard
	Co	59	8883.9	1.8	**1.2405**	0.015	1.2	ug/L	252	Standard
	Ni	60	10083.0	1.9	**5.2809**	0.102	1.9	ug/L	192	Standard
	Cu	65	17655.3	1.5	**9.2565**	0.148	1.6	ug/L	141	Standard
	Zn	66	24172.6	2.4	**21.3574**	0.493	2.3	ug/L	491	Standard
>	**Ge**	72	512038.2	0.7				ug/L	515087	Standard
	As	75	255.5	6.8	**0.2688**	0.014	5.3	ug/L	20	Standard
	Se	82	53.2	11.2	**0.3458**	0.053	15.2	ug/L	15	Standard
	Se-1	77	80.3	12.8	**0.3545**	0.144	40.6	ug/L	48	Standard
>	**Ga**	71	3018.6	3.1				mg/L	90	Standard
	Rb	85	16969.5	1.9				ug/L	23	Standard
	Y	89	529226.2	0.9				ug/L	448510	Standard
>	**Rh**	103	10.0					ug/L	15	Standard
	Mo	98	313.0	4.7	**0.0960**	0.005	5.5	ug/L	11	Standard
	Ag	107	442.0	1.6	**0.0492**	0.001	1.7	ug/L	134	Standard
	Cd	111	534.5	3.9	**0.2456**	0.011	4.4	mg/L	7	Standard
	Cd	114	1438.1	5.6	**0.2703**	0.019	7.1	ug/L	16	Standard
>	**In**	115	614980.8	1.4				ug/L	611117	Standard
	Sn	118	1005.0	9.0	**0.0670**	0.016	23.2	ug/L	518	Standard
	Sb	123	831.9	19.2	**0.1627**	0.036	21.9	ug/L	127	Standard
	Ba	135	51902.4	2.2	**24.9562**	0.305	1.2	ug/L	45	Standard
	Ce	140	320029.2	2.5				ug/L	298	Standard
>	**Tb**	159	977861.1	0.7				ug/L	977942	Standard
	Ho	165	6373.0	4.5				ug/L	12	Standard
	Tl	203	276.7	7.5	**0.0277**	0.002	8.0	ug/L	31	Standard
	Tl	205	265.0	17.0	**0.0335**	0.006	17.4	ug/L	20	Standard
	Pb	206	33506.0	0.9	**6.0646**	0.072	1.2	ug/L	286	Standard
	Pb	207	27548.0	1.0	**5.5276**	0.068	1.2	ug/L	257	Standard
	Pb	208	115638.2	1.3	**6.0000**	0.078	1.3	ug/L	956	Standard
	U	238	6967.9	1.8	**0.9584**	0.022	2.3	ug/L	17	Standard
>	**Bi**	209	563593.3	0.7				ug/L	557080	Standard

Sample ID: L1606104801
Report Date/Time: Tuesday, June 28, 2016 14:56:17
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	48.3	36.3	**0.0057**	0.024	420.3	mg/L	32	Standard
K	39	71.7	21.3	**0.0944**	0.052	54.8	mg/L	30	Standard
Ca	43	58.3	17.8	**3.2603**	1.450	44.5	mg/L	48	Standard
Fe	54	2389.6	6.6	**1.9680**	0.178	9.0	mg/L	159	Standard
Fe	57	746.7	5.6	**1.8289**	0.182	10.0	mg/L	195	Standard
Sc-1	45	40666.2	2.1				mg/L	39543	Standard
Cl	35	32405.3	1.2				ug/L	27642	Standard
Kr	83	4.0	43.3				ug/L	2	Standard
Br	81	2003.5	7.8				ug/L	1647	Standard
P	31	17084.6	3.9				ug/L	15231	Standard
S	34	2108.5	4.8				ug/L	1845	Standard
Sr	88	73.3	30.7				ug/L	70	Standard
C	12	240.0	38.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	20.0				mg/L	0	Standard
Dy	164	9760.8	1.5				mg/L	12	Standard
Ho-1	165	6373.0	4.5				mg/L	12	Standard
Er	166	5897.8	8.3				mg/L	20	Standard
I	127	2141.8	10.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		107.167	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.408	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606104801
Report Date/Time: Tuesday, June 28, 2016 14:56:17
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
In	115		100.632
Sn	118		
Sb	123		
Ba	135		
Ce	140		
Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
Bi	209		101.169
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606104801PS WG573475-03
Sample Date/Time: Tuesday, June 28, 2016 14:57:11
Number of Replicates: 3
Autosampler Position: 221
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90811.3	0.7				ug/L	81760	Standard
	Be	9	56498.5	1.8	**51.2132**	1.080	2.1	ug/L	17	Standard
	Al	27	1545775.3	0.6	**9.0143**	0.017	0.2	ug/L	1410	Standard
	Sc	45	41961.4	1.6				ug/L	39543	Standard
	Ti	47	342.0	11.2	**1.1864**	0.157	13.2	ug/L	39	Standard
	V	51	392129.2	0.5	**55.2170**	0.488	0.9	ug/L	1415	Standard
	Cr	52	384433.3	0.4	**53.1980**	0.445	0.8	ug/L	6946	Standard
	Cr	53	48853.7	2.8	**54.3367**	1.767	3.3	ug/L	328	Standard
	Mn	55	478138.1	0.8	**72.3045**	0.492	0.7	ug/L	823	Standard
	Co	59	389747.9	0.3	**54.4630**	0.290	0.5	ug/L	252	Standard
	Ni	60	106262.0	0.3	**55.3242**	0.400	0.7	ug/L	192	Standard
	Cu	65	114943.7	0.4	**59.3369**	0.368	0.6	ug/L	141	Standard
	Zn	66	80976.1	0.1	**71.7501**	0.396	0.6	ug/L	491	Standard
>	**Ge**	72	524333.4	0.5				ug/L	515087	Standard
	As	75	59243.0	1.9	**51.8077**	1.242	2.4	ug/L	20	Standard
	Se	82	6211.1	2.1	**52.8798**	1.213	2.3	ug/L	15	Standard
	Se-1	77	4085.6	1.2	**52.5558**	0.559	1.1	ug/L	48	Standard
>	**Ga**	71	3287.0	1.8				mg/L	90	Standard
	Rb	85	16744.3	3.7				ug/L	23	Standard
	Y	89	522998.5	1.8				ug/L	448510	Standard
>	**Rh**	103	28.3	36.7				ug/L	15	Standard
	Mo	98	332.0	9.1	**0.1017**	0.011	11.1	ug/L	11	Standard
	Ag	107	350941.4	1.0	**51.3750**	0.335	0.7	ug/L	134	Standard
	Cd	111	110107.7	1.0	**50.6313**	0.348	0.7	mg/L	7	Standard
	Cd	114	268982.1	1.5	**50.8849**	0.677	1.3	ug/L	16	Standard
>	**In**	115	621693.0	0.4				ug/L	611117	Standard
	Sn	118	1258.4	5.6	**0.1073**	0.012	10.9	ug/L	518	Standard
	Sb	123	241746.4	1.3	**50.3512**	0.498	1.0	ug/L	127	Standard
	Ba	135	155747.4	0.9	**74.1325**	0.655	0.9	ug/L	45	Standard
	Ce	140	312396.1	0.8				ug/L	298	Standard
>	**Tb**	159	997317.5	1.6				ug/L	977942	Standard
	Ho	165	6548.1	1.9				ug/L	12	Standard
	Tl	203	447292.7	0.3	**50.4657**	0.615	1.2	ug/L	31	Standard
	Tl	205	398193.6	1.5	**51.7351**	0.805	1.6	ug/L	20	Standard
	Pb	206	325033.3	0.1	**57.1116**	0.488	0.9	ug/L	286	Standard
	Pb	207	282464.7	0.8	**55.0228**	0.100	0.2	ug/L	257	Standard
	Pb	208	1144735.7	1.0	**57.6431**	0.938	1.6	ug/L	956	Standard
	U	238	386201.4	1.1	**51.2415**	0.797	1.6	ug/L	17	Standard
>	**Bi**	209	586004.9	0.9				ug/L	557080	Standard

Sample ID: L1606104801PS WG573475-03
Report Date/Time: Tuesday, June 28, 2016 14:59:28
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	48.3	41.8	**0.0038**	0.026	698.9	mg/L	32	Standard
K	39	65.0	20.4	**0.0660**	0.045	68.1	mg/L	30	Standard
Ca	43	46.7	27.0	**1.3158**	1.694	128.8	mg/L	48	Standard
Fe	54	2337.7	11.4	**1.8584**	0.246	13.3	mg/L	159	Standard
Fe	57	760.0	13.2	**1.7916**	0.300	16.7	mg/L	195	Standard
> **Sc-1**	45	41961.4	1.6				mg/L	39543	Standard
Cl	35	32737.3	1.7				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	2180.2	6.6				ug/L	1647	Standard
P	31	17061.3	3.2				ug/L	15231	Standard
S	34	2245.2	3.3				ug/L	1845	Standard
Sr	88	81.7	12.7				ug/L	70	Standard
C	12	226.7	6.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	46.7	32.7				mg/L	0	Standard
Dy	164	10105.0	3.1				mg/L	12	Standard
Ho-1	165	6548.1	1.9				mg/L	12	Standard
Er	166	5884.5	3.8				mg/L	20	Standard
I	127	2158.5	15.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		111.071	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		101.795	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606104801PS WG573475-03
Report Date/Time: Tuesday, June 28, 2016 14:59:28
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.731
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.192
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606104801SDL WG573475-04
Sample Date/Time: Tuesday, June 28, 2016 15:00:23
Number of Replicates: 3
Autosampler Position: 222
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89188.5	2.4				ug/L	81760	Standard
	Be	9	71.7	4.0	**0.0564**	0.001	2.3	ug/L	17	Standard
	Al	27	302683.6	0.7	**1.7866**	0.056	3.1	ug/L	1410	Standard
	Sc	45	41125.8	3.4				ug/L	39543	Standard
	Ti	47	82.7	6.2	**0.1644**	0.030	18.2	ug/L	39	Standard
	V	51	6527.8	1.5	**0.7233**	0.030	4.1	ug/L	1415	Standard
	Cr	52	10512.6	0.6	**0.4384**	0.054	12.4	ug/L	6946	Standard
	Cr	53	805.0	22.0	**0.5666**	0.199	35.2	ug/L	328	Standard
	Mn	55	27749.4	1.2	**4.0892**	0.112	2.7	ug/L	823	Standard
	Co	59	1936.1	6.5	**0.2412**	0.012	4.8	ug/L	252	Standard
	Ni	60	2147.2	1.2	**1.0241**	0.047	4.6	ug/L	192	Standard
	Cu	65	3398.4	1.3	**1.6803**	0.070	4.1	ug/L	141	Standard
	Zn	66	7011.6	1.2	**5.5054**	0.150	2.7	ug/L	491	Standard
>	**Ge**	72	520536.2	3.0				ug/L	515087	Standard
	As	75	60.9	66.9	**0.0932**	0.035	37.4	ug/L	20	Standard
	Se	82	24.2	49.2	**0.0869**	0.099	113.8	ug/L	15	Standard
	Se-1	77	55.3	11.8	**0.0096**	0.103	1081.5	ug/L	48	Standard
>	**Ga**	71	663.3	7.6				mg/L	90	Standard
	Rb	85	3270.4	5.2				ug/L	23	Standard
	Y	89	464717.5	1.9				ug/L	448510	Standard
>	**Rh**	103	16.7	121.2				ug/L	15	Standard
	Mo	98	80.8	3.1	**0.0096**	0.001	7.0	ug/L	11	Standard
	Ag	107	180.0	20.5	**0.0094**	0.005	54.7	ug/L	134	Standard
	Cd	111	113.9	5.5	**0.0481**	0.003	5.7	mg/L	7	Standard
	Cd	114	303.5	12.8	**0.0510**	0.007	13.8	ug/L	16	Standard
>	**In**	115	638226.3	1.3				ug/L	611117	Standard
	Sn	118	801.7	13.2	**0.0279**	0.017	59.8	ug/L	518	Standard
	Sb	123	2112.7	26.2	**0.4158**	0.111	26.6	ug/L	127	Standard
	Ba	135	10186.4	1.4	**4.7001**	0.102	2.2	ug/L	45	Standard
	Ce	140	62112.1	0.6				ug/L	298	Standard
>	**Tb**	159	1019737.0	1.0				ug/L	977942	Standard
	Ho	165	1263.4	4.0				ug/L	12	Standard
	Tl	203	113.0	16.9	**0.0078**	0.002	24.0	ug/L	31	Standard
	Tl	205	118.3	30.6	**0.0128**	0.004	33.6	ug/L	20	Standard
	Pb	206	6901.2	1.0	**1.1348**	0.029	2.6	ug/L	286	Standard
	Pb	207	5663.1	0.8	**1.0321**	0.019	1.9	ug/L	257	Standard
	Pb	208	22745.7	0.6	**1.0713**	0.025	2.3	ug/L	956	Standard
	U	238	1347.1	2.4	**0.1734**	0.003	1.5	ug/L	17	Standard
>	**Bi**	209	593755.8	1.9				ug/L	557080	Standard

Sample ID: L1606104801SDL WG573475-04
Report Date/Time: Tuesday, June 28, 2016 15:02:40
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	45.0	40.1	**0.0009**	0.025	2626.8	mg/L	32	Standard
K	39	30.0	28.9	**-0.0443**	0.026	58.3	mg/L	30	Standard
Ca	43	28.3	20.4	**-1.2169**	0.752	61.8	mg/L	48	Standard
Fe	54	584.0	13.3	**0.3714**	0.068	18.2	mg/L	159	Standard
Fe	57	295.0	14.5	**0.3314**	0.118	35.5	mg/L	195	Standard
> **Sc-1**	45	41125.8	3.4				mg/L	39543	Standard
Cl	35	31826.1	1.1				ug/L	27642	Standard
Kr	83	2.7	21.7				ug/L	2	Standard
Br	81	1903.5	3.4				ug/L	1647	Standard
P	31	16692.5	3.1				ug/L	15231	Standard
S	34	2128.5	4.6				ug/L	1845	Standard
Sr	88	68.3	36.1				ug/L	70	Standard
C	12	176.7	42.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	1953.3	7.3				mg/L	12	Standard
Ho-1	165	1263.4	4.0				mg/L	12	Standard
Er	166	1116.7	6.8				mg/L	20	Standard
I	127	1450.1	11.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		109.086	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.058	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606104801SDL WG573475-04
Report Date/Time: Tuesday, June 28, 2016 15:02:40
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.436
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.584
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Tuesday, June 28, 2016 15:03:36
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	89282.3	1.4				ug/L	81760	Standard
	Be	9	15.0	33.3	**0.0041**	0.005	114.8	ug/L	17	Standard
	Al	27	7897505.8	1.4	**46.9037**	0.018	0.0	ug/L	1410	Standard
	Sc	45	39757.1	3.6				ug/L	39543	Standard
	Ti	47	27232.5	2.5	**109.9775**	3.234	2.9	ug/L	39	Standard
	V	51	789.3	35.9	**-0.0922**	0.042	45.6	ug/L	1415	Standard
	Cr	52	7275.7	2.2	**-0.0059**	0.028	478.1	ug/L	6946	Standard
	Cr	53	6823.2	10.4	**7.4758**	0.819	11.0	ug/L	328	Standard
	Mn	55	1397.7	2.4	**0.0683**	0.004	5.9	ug/L	823	Standard
	Co	59	388.0	3.6	**0.0242**	0.002	7.5	ug/L	252	Standard
	Ni	60	839.0	6.3	**0.3431**	0.028	8.3	ug/L	192	Standard
	Cu	65	614.0	0.9	**0.2341**	0.001	0.3	ug/L	141	Standard
	Zn	66	1957.1	1.8	**0.9682**	0.037	3.8	ug/L	491	Standard
>	Ge	72	512179.1	1.0				ug/L	515087	Standard
	As	75	-64.0	70.1	**-0.0170**	0.040	237.0	ug/L	20	Standard
	Se	82	21.9	13.2	**0.0720**	0.027	37.3	ug/L	15	Standard
	Se-1	77	397.7	7.2	**4.5923**	0.428	9.3	ug/L	48	Standard
>	Ga	71	168.3	7.5				mg/L	90	Standard
	Rb	85	1135.0	8.9				ug/L	23	Standard
	Y	89	442720.4	1.1				ug/L	448510	Standard
>	Rh	103	10.0	50.0				ug/L	15	Standard
	Mo	98	272741.3	0.7	**100.8481**	1.185	1.2	ug/L	11	Standard
	Ag	107	262.0	2.0	**0.0228**	0.001	4.1	ug/L	134	Standard
	Cd	111	-198.2	9.3	**-0.0956**	0.008	8.7	mg/L	7	Standard
	Cd	114	465.1	17.4	**0.0845**	0.015	17.7	ug/L	16	Standard
>	In	115	611378.2	0.9				ug/L	611117	Standard
	Sn	118	743.4	13.4	**0.0238**	0.017	73.3	ug/L	518	Standard
	Sb	123	656.7	25.2	**0.1264**	0.036	28.5	ug/L	127	Standard
	Ba	135	610.7	6.5	**0.2705**	0.017	6.2	ug/L	45	Standard
	Ce	140	266.7	9.2				ug/L	298	Standard
>	Tb	159	977869.4	2.5				ug/L	977942	Standard
	Ho	165	16.7	17.3				ug/L	12	Standard
	Tl	203	193.3	8.8	**0.0182**	0.002	10.2	ug/L	31	Standard
	Tl	205	168.3	10.4	**0.0207**	0.002	10.8	ug/L	20	Standard
	Pb	206	552.7	5.7	**0.0387**	0.005	12.2	ug/L	286	Standard
	Pb	207	474.0	2.6	**0.0392**	0.001	3.3	ug/L	257	Standard
	Pb	208	1934.7	0.9	**0.0422**	0.002	5.1	ug/L	956	Standard
	U	238	35.3	29.7	**0.0020**	0.001	71.8	ug/L	17	Standard
>	Bi	209	556811.8	1.3				ug/L	557080	Standard

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 15:05:53
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**20.5038**	35.505	173.2	mg/L	0	Standard
Mg	24	10406.9	1.4	**13.7318**	0.472	3.4	mg/L	32	Standard
K	39	1353.4	4.9	**4.4093**	0.243	5.5	mg/L	30	Standard
Ca	43	183.3	3.1	**22.3899**	1.779	7.9	mg/L	48	Standard
Fe	54	14710.0	1.7	**13.1205**	0.693	5.3	mg/L	159	Standard
Fe	57	4182.2	3.4	**13.4196**	0.181	1.3	mg/L	195	Standard
> **Sc-1**	45	39757.1	3.6				mg/L	39543	Standard
Cl	35	33087.4	2.4				ug/L	27642	Standard
Kr	83	1.3	173.2				ug/L	2	Standard
Br	81	1766.8	6.4				ug/L	1647	Standard
P	31	17728.7	3.7				ug/L	15231	Standard
S	34	2340.2	3.3				ug/L	1845	Standard
Sr	88	45.0	0.0				ug/L	70	Standard
C	12	306.7	6.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	93.3	12.4				mg/L	0	Standard
Dy	164	15.1	45.8				mg/L	12	Standard
Ho-1	165	16.7	17.3				mg/L	12	Standard
Er	166	33.3	75.5				mg/L	20	Standard
I	127	1891.8	6.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27	0.938		
Sc	45			
Ti	47	109.978		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.435	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 15:05:53
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	100.848		
Ag	107			
Cd	111			
Cd	114			
In	115			100.043
Sn	118			
Sb	123			
Ba	135			
Ce	140			
Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
Bi	209			99.952
Na	23	164.030		
Mg	24	274.635		
K	39	88.187		
Ca	43	149.266		
Fe	54	104.964		
Fe	57	107.357		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Na	23	
QC Std 4	Mg	24	

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 15:05:53
Page 3

Approved: June 29, 2016

QC Std 4	Ca	43

Sample ID: QC Std 4
Report Date/Time: Tuesday, June 28, 2016 15:05:53
Page 4

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Tuesday, June 28, 2016 15:06:47
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	87968.0	1.5				ug/L	81760	Standard
	Be	9	111223.4	1.1	**104.1090**	2.678	2.6	ug/L	17	Standard
	Al	27	8303576.6	1.0	**50.0632**	1.083	2.2	ug/L	1410	Standard
	Sc	45	38970.1	2.5				ug/L	39543	Standard
	Ti	47	28030.6	1.5	**112.6949**	1.093	1.0	ug/L	39	Standard
	V	51	696349.8	1.4	**100.1131**	1.784	1.8	ug/L	1415	Standard
	Cr	52	683707.3	0.8	**97.2904**	0.687	0.7	ug/L	6946	Standard
	Cr	53	91713.5	1.8	**104.2725**	1.761	1.7	ug/L	328	Standard
	Mn	55	633248.9	0.6	**97.6578**	0.701	0.7	ug/L	823	Standard
	Co	59	716430.9	0.3	**102.0719**	1.125	1.1	ug/L	252	Standard
	Ni	60	180604.4	0.8	**95.9232**	1.625	1.7	ug/L	192	Standard
	Cu	65	185052.2	0.7	**97.4326**	1.562	1.6	ug/L	141	Standard
	Zn	66	113502.2	0.3	**102.8618**	0.650	0.6	ug/L	491	Standard
>	Ge	72	514438.0	0.9				ug/L	515087	Standard
	As	75	114188.1	0.3	**101.7361**	0.865	0.8	ug/L	20	Standard
	Se	82	11752.5	0.4	**102.0988**	1.402	1.4	ug/L	15	Standard
	Se-1	77	8079.2	0.7	**106.6610**	0.727	0.7	ug/L	48	Standard
>	Ga	71	190.0	7.0				mg/L	90	Standard
	Rb	85	893.4	6.0				ug/L	23	Standard
	Y	89	434662.6	1.6				ug/L	448510	Standard
>	Rh	103	38.3	52.7				ug/L	15	Standard
	Mo	98	282406.5	1.5	**107.3339**	1.537	1.4	ug/L	11	Standard
	Ag	107	594345.7	0.8	**90.9666**	1.393	1.5	ug/L	134	Standard
	Cd	111	207638.0	0.2	**99.8149**	1.587	1.6	mg/L	7	Standard
	Cd	114	499764.2	0.7	**98.8323**	1.071	1.1	ug/L	16	Standard
>	In	115	594810.8	1.6				ug/L	611117	Standard
	Sn	118	1138.4	7.8	**0.0959**	0.015	15.6	ug/L	518	Standard
	Sb	123	463528.1	1.1	**100.9493**	2.612	2.6	ug/L	127	Standard
	Ba	135	197524.5	0.1	**98.2892**	1.431	1.5	ug/L	45	Standard
	Ce	140	85.0	21.2				ug/L	298	Standard
>	Tb	159	967938.8	1.8				ug/L	977942	Standard
	Ho	165	8.3	69.3				ug/L	12	Standard
	Tl	203	848553.4	0.8	**99.1555**	1.255	1.3	ug/L	31	Standard
	Tl	205	762572.5	0.3	**102.6175**	1.400	1.4	ug/L	20	Standard
	Pb	206	559247.5	0.6	**101.8213**	1.258	1.2	ug/L	286	Standard
	Pb	207	488004.9	0.6	**98.5027**	1.226	1.2	ug/L	257	Standard
	Pb	208	1955600.4	0.4	**102.0354**	1.641	1.6	ug/L	956	Standard
	U	238	737824.6	0.7	**101.3974**	2.083	2.1	ug/L	17	Standard
>	Bi	209	565855.3	1.5				ug/L	557080	Standard

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 15:09:04
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**21.8863**	18.964	86.6	mg/L	0	Standard
Mg	24	10742.1	1.4	**14.4601**	0.490	3.4	mg/L	32	Standard
K	39	1261.7	7.3	**4.1815**	0.232	5.5	mg/L	30	Standard
Ca	43	183.3	29.9	**22.9918**	8.781	38.2	mg/L	48	Standard
Fe	54	15164.6	1.1	**13.7950**	0.409	3.0	mg/L	159	Standard
Fe	57	4003.9	4.1	**13.0952**	0.606	4.6	mg/L	195	Standard
Sc-1	45	38970.1	2.5				mg/L	39543	Standard
Cl	35	31476.7	2.0				ug/L	27642	Standard
Kr	83	3.3	75.5				ug/L	2	Standard
Br	81	1726.8	7.1				ug/L	1647	Standard
P	31	16804.3	1.0				ug/L	15231	Standard
S	34	2221.8	2.4				ug/L	1845	Standard
Sr	88	46.7	22.3				ug/L	70	Standard
C	12	280.0	21.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	0	Standard
Dy	164	25.2	45.9				mg/L	12	Standard
Ho-1	165	8.3	69.3				mg/L	12	Standard
Er	166	30.0	66.7				mg/L	20	Standard
I	127	1238.4	19.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	104.109		
Al	27	1.001		
Sc	45			
Ti	47	112.695		
V	51	100.113		
Cr	52	97.290		
Cr	53			
Mn	55	97.658		
Co	59	102.072		
Ni	60	95.923		
Cu	65	97.433		
Zn	66	102.862		
> Ge	72		99.874	
As	75	101.736		
Se	82	102.099		
Se-1	77			
> Ga	71			

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 15:09:04
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	107.334		
Ag	107	90.967		
Cd	111	99.815		
Cd	114			
In	115		97.332	
Sn	118			
Sb	123	100.949		
Ba	135	98.289		
Ce	140			
Tb	159			
Ho	165			
Tl	203	99.155		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.035		
U	238	101.397		
Bi	209		101.575	
Na	23	175.090		
Mg	24	289.203		
K	39	83.630		
Ca	43	153.278		
Fe	54	110.360		
Fe	57	104.761		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Na	23	
QC Std 5	Mg	24	

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 15:09:04
Page 3

Approved: June 29, 2016

QC Std 5	Ca	43

Sample ID: QC Std 5
Report Date/Time: Tuesday, June 28, 2016 15:09:04
Page 4

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 15:10:01
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85172.2	3.3				ug/L	81760	Standard
	Be	9	53236.8	4.1	**51.4767**	2.435	4.7	ug/L	17	Standard
	Al	27	7628795.8	0.6	**47.5256**	1.416	3.0	ug/L	1410	Standard
	Sc	45	37696.8	2.0				ug/L	39543	Standard
	Ti	47	24164.9	0.9	**98.0401**	2.312	2.4	ug/L	39	Standard
	V	51	343835.2	0.1	**49.7831**	0.751	1.5	ug/L	1415	Standard
	Cr	52	344282.1	1.0	**48.9253**	1.054	2.2	ug/L	6946	Standard
	Cr	53	43472.3	1.9	**49.7131**	1.693	3.4	ug/L	328	Standard
	Mn	55	313567.9	1.4	**48.7360**	1.405	2.9	ug/L	823	Standard
	Co	59	359786.6	1.1	**51.7171**	1.082	2.1	ug/L	252	Standard
	Ni	60	90351.3	0.3	**48.3759**	0.827	1.7	ug/L	192	Standard
	Cu	65	92251.0	0.7	**48.9699**	0.708	1.4	ug/L	141	Standard
	Zn	66	54515.2	1.1	**49.4372**	1.191	2.4	ug/L	491	Standard
>	**Ge**	72	509796.1	1.5				ug/L	515087	Standard
	As	75	55417.4	0.8	**49.8530**	1.118	2.2	ug/L	20	Standard
	Se	82	5885.8	2.0	**51.5527**	1.791	3.5	ug/L	15	Standard
	Se-1	77	3913.8	2.3	**51.7920**	1.979	3.8	ug/L	48	Standard
>	**Ga**	71	60.0	14.4				mg/L	90	Standard
	Rb	85	760.0	3.9				ug/L	23	Standard
	Y	89	437496.8	1.0				ug/L	448510	Standard
>	**Rh**	103	35.0	28.6				ug/L	15	Standard
	Mo	98	271668.8	0.6	**101.2993**	0.747	0.7	ug/L	11	Standard
	Ag	107	329448.9	0.2	**49.4598**	0.339	0.7	ug/L	134	Standard
	Cd	111	103364.2	1.7	**48.7409**	0.531	1.1	mg/L	7	Standard
	Cd	114	251263.5	1.7	**48.7432**	0.553	1.1	ug/L	16	Standard
>	**In**	115	606236.0	0.6				ug/L	611117	Standard
	Sn	118	288784.4	0.3	**49.1320**	0.341	0.7	ug/L	518	Standard
	Sb	123	235993.4	0.4	**50.4089**	0.456	0.9	ug/L	127	Standard
	Ba	135	99324.5	0.4	**48.4735**	0.237	0.5	ug/L	45	Standard
	Ce	140	126.7	35.6				ug/L	298	Standard
>	**Tb**	159	959991.8	0.1				ug/L	977942	Standard
	Ho	165	8.3	34.6				ug/L	12	Standard
	Tl	203	427212.2	0.6	**50.5774**	0.492	1.0	ug/L	31	Standard
	Tl	205	380951.0	2.1	**51.9422**	1.422	2.7	ug/L	20	Standard
	Pb	206	273119.7	0.4	**50.3511**	0.572	1.1	ug/L	286	Standard
	Pb	207	248886.2	0.4	**50.8720**	0.472	0.9	ug/L	257	Standard
	Pb	208	972505.1	1.3	**51.3811**	0.993	1.9	ug/L	956	Standard
	U	238	367751.0	1.4	**51.2030**	1.055	2.1	ug/L	17	Standard
>	**Bi**	209	558440.3	0.7				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:12:17
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4133.9	5.5	**5.7164**	0.326	5.7	mg/L	32	Standard
K	39	1253.4	8.2	**4.3011**	0.352	8.2	mg/L	30	Standard
Ca	43	70.0	37.8	**5.7557**	4.052	70.4	mg/L	48	Standard
Fe	54	5742.0	2.6	**5.3173**	0.253	4.8	mg/L	159	Standard
Fe	57	1578.4	3.1	**4.9652**	0.143	2.9	mg/L	195	Standard
> **Sc-1**	45	37696.8	2.0				mg/L	39543	Standard
Cl	35	31580.2	0.9				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	1660.1	9.6				ug/L	1647	Standard
P	31	16826.0	2.3				ug/L	15231	Standard
S	34	2313.5	5.0				ug/L	1845	Standard
Sr	88	61.7	26.1				ug/L	70	Standard
C	12	133.3	18.9				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	13.3	114.6				mg/L	12	Standard
Ho-1	165	8.3	34.6				mg/L	12	Standard
Er	166	0.0					mg/L	20	Standard
I	127	1290.1	24.6				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	102.953		
	Al	27	95.051		
	Sc	45			
	Ti	47	98.040		
	V	51	99.566		
	Cr	52	97.851		
	Cr	53			
	Mn	55	97.472		
	Co	59	103.434		
	Ni	60	96.752		
	Cu	65	97.940		
	Zn	66	98.874		
>	Ge	72		98.973	
	As	75	99.706		
	Se	82	103.105		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:12:17
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	101.299		
Ag	107	98.920		
Cd	111	97.482		
Cd	114			
> In	115			99.201
Sn	118	98.264		
Sb	123	100.818		
Ba	135	96.947		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.155		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.762		
U	238	102.406		
> Bi	209			100.244
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:12:17
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 15:13:12
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	88670.5	2.8				ug/L	81760	Standard
	Be	9	51.7	24.4	**0.0381**	0.010	27.4	ug/L	17	Standard
	Al	27	5021.0	57.4	**0.0153**	0.016	105.4	ug/L	1410	Standard
	Sc	45	39924.2	3.2				ug/L	39543	Standard
	Ti	47	36.7	33.9	**-0.0173**	0.048	276.4	ug/L	39	Standard
	V	51	1293.4	14.6	**-0.0198**	0.025	126.5	ug/L	1415	Standard
	Cr	52	6153.2	0.9	**-0.1686**	0.020	12.1	ug/L	6946	Standard
	Cr	53	480.0	7.9	**0.2078**	0.038	18.2	ug/L	328	Standard
	Mn	55	818.0	20.7	**-0.0219**	0.025	112.7	ug/L	823	Standard
	Co	59	362.0	35.7	**0.0203**	0.018	87.2	ug/L	252	Standard
	Ni	60	216.7	22.5	**0.0105**	0.024	233.6	ug/L	192	Standard
	Cu	65	206.0	48.8	**0.0177**	0.051	290.8	ug/L	141	Standard
	Zn	66	512.3	23.8	**-0.3586**	0.106	29.6	ug/L	491	Standard
>	Ge	72	512537.5	1.6				ug/L	515087	Standard
	As	75	16.6	529.1	**0.0544**	0.078	142.5	ug/L	20	Standard
	Se	82	20.8	66.0	**0.0610**	0.117	191.3	ug/L	15	Standard
	Se-1	77	56.0	20.1	**0.0270**	0.140	518.0	ug/L	48	Standard
>	Ga	71	48.3	15.8				mg/L	90	Standard
	Rb	85	23.3	32.7				ug/L	23	Standard
	Y	89	442158.0	1.4				ug/L	448510	Standard
>	Rh	103	8.3	69.3				ug/L	15	Standard
	Mo	98	390.5	34.5	**0.1230**	0.050	40.8	ug/L	11	Standard
	Ag	107	293.7	71.4	**0.0269**	0.031	115.5	ug/L	134	Standard
	Cd	111	93.5	118.9	**0.0403**	0.052	128.0	mg/L	7	Standard
	Cd	114	191.9	127.3	**0.0316**	0.047	147.6	ug/L	16	Standard
>	In	115	622178.9	0.7				ug/L	611117	Standard
	Sn	118	798.4	19.1	**0.0306**	0.025	81.3	ug/L	518	Standard
	Sb	123	1327.2	22.5	**0.2633**	0.061	23.3	ug/L	127	Standard
	Ba	135	121.3	95.4	**0.0329**	0.056	169.3	ug/L	45	Standard
	Ce	140	156.7	109.9				ug/L	298	Standard
>	Tb	159	979098.8	1.1				ug/L	977942	Standard
	Ho	165	8.3	124.9				ug/L	12	Standard
	Tl	203	319.0	100.9	**0.0316**	0.036	115.5	ug/L	31	Standard
	Tl	205	186.7	86.1	**0.0222**	0.021	94.2	ug/L	20	Standard
	Pb	206	461.3	35.4	**0.0186**	0.029	153.9	ug/L	286	Standard
	Pb	207	392.3	40.1	**0.0194**	0.031	157.4	ug/L	257	Standard
	Pb	208	1505.0	29.2	**0.0164**	0.022	134.2	ug/L	956	Standard
	U	238	157.3	69.9	**0.0181**	0.015	80.5	ug/L	17	Standard
>	Bi	209	578116.8	1.4				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:15:29
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	31.7	18.2	**-0.0152**	0.009	57.6	mg/L	32	Standard
K	39	23.3	49.5	**-0.0625**	0.040	64.6	mg/L	30	Standard
Ca	43	25.0	40.0	**-1.6032**	1.411	88.0	mg/L	48	Standard
Fe	54	127.6	20.0	**-0.0223**	0.025	111.2	mg/L	159	Standard
Fe	57	173.3	7.3	**-0.0450**	0.038	85.1	mg/L	195	Standard
Sc-1	45	39924.2	3.2				mg/L	39543	Standard
Cl	35	32496.1	0.6				ug/L	27642	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	1573.4	6.5				ug/L	1647	Standard
P	31	17568.5	2.8				ug/L	15231	Standard
S	34	2373.5	3.0				ug/L	1845	Standard
Sr	88	50.0	26.5				ug/L	70	Standard
C	12	106.7	19.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	12.5	38.4				mg/L	12	Standard
Ho-1	165	8.3	124.9				mg/L	12	Standard
Er	166	16.7	124.9				mg/L	20	Standard
I	127	1251.7	20.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.505	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:15:29
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.810
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.776
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Sb	123	

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:15:29
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Tuesday, June 28, 2016 15:16:29
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81011.1	2.2				ug/L	81760	Standard
	Be	9	195.0	20.5	**0.1880**	0.037	19.7	ug/L	17	Standard
	Al	27	1493.4	21.7	**-0.0047**	0.002	41.0	ug/L	1410	Standard
	Sc	45	36542.4	1.6				ug/L	39543	Standard
	Ti	47	30.7	9.4	**-0.0369**	0.011	30.3	ug/L	39	Standard
	V	51	3567.1	2.8	**0.3274**	0.017	5.3	ug/L	1415	Standard
	Cr	52	11052.7	0.9	**0.5951**	0.027	4.6	ug/L	6946	Standard
	Cr	53	1095.0	4.8	**0.9563**	0.056	5.9	ug/L	328	Standard
	Mn	55	3466.7	0.6	**0.4079**	0.007	1.6	ug/L	823	Standard
	Co	59	2834.6	2.1	**0.3884**	0.006	1.6	ug/L	252	Standard
	Ni	60	2871.9	0.5	**1.4818**	0.019	1.3	ug/L	192	Standard
	Cu	65	1515.4	2.2	**0.7391**	0.023	3.1	ug/L	141	Standard
	Zn	66	7448.8	1.9	**6.2435**	0.137	2.2	ug/L	491	Standard
>	**Ge**	72	495055.6	0.7				ug/L	515087	Standard
	As	75	363.8	8.2	**0.3769**	0.025	6.6	ug/L	20	Standard
	Se	82	58.7	3.6	**0.4109**	0.015	3.7	ug/L	15	Standard
	Se-1	77	71.3	17.1	**0.2665**	0.171	64.2	ug/L	48	Standard
>	**Ga**	71	38.3	19.9				mg/L	90	Standard
	Rb	85	26.7	28.6				ug/L	23	Standard
	Y	89	415588.9	2.9				ug/L	448510	Standard
>	**Rh**	103	6.7	43.3				ug/L	15	Standard
	Mo	98	101.2	14.2	**0.0199**	0.005	26.4	ug/L	11	Standard
	Ag	107	2581.2	2.0	**0.3840**	0.010	2.6	ug/L	134	Standard
	Cd	111	489.5	5.9	**0.2354**	0.013	5.3	mg/L	7	Standard
	Cd	114	1118.1	7.7	**0.2189**	0.015	6.9	ug/L	16	Standard
>	**In**	115	587241.6	1.2				ug/L	611117	Standard
	Sn	118	508.3	8.2	**-0.0124**	0.008	65.2	ug/L	518	Standard
	Sb	123	2001.1	5.7	**0.4288**	0.031	7.2	ug/L	127	Standard
	Ba	135	1392.4	6.1	**0.6767**	0.039	5.8	ug/L	45	Standard
	Ce	140	30.0	44.1				ug/L	298	Standard
>	**Tb**	159	929212.6	0.4				ug/L	977942	Standard
	Ho	165	13.3	57.3				ug/L	12	Standard
	Tl	203	711.0	4.8	**0.0804**	0.004	5.1	ug/L	31	Standard
	Tl	205	656.7	5.8	**0.0883**	0.005	5.1	ug/L	20	Standard
	Pb	206	1347.1	1.6	**0.1880**	0.004	2.0	ug/L	286	Standard
	Pb	207	1117.4	2.9	**0.1733**	0.005	3.1	ug/L	257	Standard
	Pb	208	4713.3	2.2	**0.1919**	0.007	3.7	ug/L	956	Standard
	U	238	2843.3	0.8	**0.3977**	0.005	1.3	ug/L	17	Standard
>	**Bi**	209	551791.7	0.8				ug/L	557080	Standard

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 15:18:46
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	26.7	39.0	**-0.0189**	0.015	77.3	mg/L	32	Standard
K	39	20.0	25.0	**-0.0684**	0.018	25.7	mg/L	30	Standard
Ca	43	25.0	40.0	**-1.2235**	1.703	139.2	mg/L	48	Standard
Fe	54	112.3	6.9	**-0.0270**	0.008	31.0	mg/L	159	Standard
Fe	57	200.0	28.8	**0.1050**	0.205	195.5	mg/L	195	Standard
Sc-1	45	36542.4	1.6				mg/L	39543	Standard
Cl	35	31755.9	1.1				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1523.4	8.9				ug/L	1647	Standard
P	31	12263.3	0.6				ug/L	15231	Standard
S	34	2283.5	2.6				ug/L	1845	Standard
Sr	88	63.3	25.4				ug/L	70	Standard
C	12	146.7	38.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	15.9	37.5				mg/L	12	Standard
Ho-1	165	13.3	57.3				mg/L	12	Standard
Er	166	16.7	91.7				mg/L	20	Standard
I	127	1103.4	23.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	94.012		
Al	27			
Sc	45			
Ti	47			
V	51	81.855		
Cr	52	74.385		
Cr	53			
Mn	55	81.586		
Co	59	97.096		
Ni	60	92.611		
Cu	65	92.383		
Zn	66	99.896		
Ge	72		96.111	
As	75	94.213		
Se	82	102.730		
Se-1	77			
Ga	71			

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 15:18:46
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	96.001		
Cd	111	98.068		
Cd	114			
> In	115			96.093
Sn	118			
Sb	123	107.193		
Ba	135	90.223		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	100.519		
Tl	205			
Pb	206			
Pb	207			
Pb	208	95.932		
U	238	99.422		
> Bi	209			99.051
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 15:18:46
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078703 WG573333-01
Sample Date/Time: Tuesday, June 28, 2016 15:24:24
Number of Replicates: 3
Autosampler Position: 223
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77973.6	3.5				ug/L	81760	Standard
	Be	9	93.3	32.7	**0.0882**	0.029	33.1	ug/L	17	Standard
	Al	27	522182.4	1.6	**3.5410**	0.149	4.2	ug/L	1410	Standard
	Sc	45	36268.4	3.8				ug/L	39543	Standard
	Ti	47	171.7	14.8	**0.5849**	0.118	20.2	ug/L	39	Standard
	V	51	8345.2	3.6	**1.0965**	0.041	3.8	ug/L	1415	Standard
	Cr	52	9135.4	1.3	**0.3675**	0.032	8.6	ug/L	6946	Standard
	Cr	53	715.0	6.0	**0.5424**	0.045	8.2	ug/L	328	Standard
	Mn	55	230272.3	2.4	**38.4029**	0.677	1.8	ug/L	823	Standard
	Co	59	4116.2	4.4	**0.6044**	0.024	3.9	ug/L	252	Standard
	Ni	60	2251.2	4.1	**1.1925**	0.054	4.6	ug/L	192	Standard
	Cu	65	1784.1	4.6	**0.9285**	0.051	5.5	ug/L	141	Standard
	Zn	66	3508.4	21.7	**2.6509**	0.786	29.6	ug/L	491	Standard
>	**Ge**	72	474556.6	0.9				ug/L	515087	Standard
	As	75	172.3	13.5	**0.2068**	0.024	11.5	ug/L	20	Standard
	Se	82	19.4	8.6	**0.0638**	0.017	27.3	ug/L	15	Standard
	Se-1	77	47.7	3.2	**-0.0320**	0.028	88.4	ug/L	48	Standard
>	**Ga**	71	883.4	11.6				mg/L	90	Standard
	Rb	85	7710.3	5.5				ug/L	23	Standard
	Y	89	442484.5	1.8				ug/L	448510	Standard
>	**Rh**	103	20.0	25.0				ug/L	15	Standard
	Mo	98	50.0	20.9	**0.0005**	0.004	864.7	ug/L	11	Standard
	Ag	107	103.7	35.8	**0.0001**	0.006	5294.7	ug/L	134	Standard
	Cd	111	33.9	43.4	**0.0139**	0.007	52.6	mg/L	7	Standard
	Cd	114	75.2	57.1	**0.0104**	0.009	84.2	ug/L	16	Standard
>	**In**	115	577946.8	0.2				ug/L	611117	Standard
	Sn	118	366.7	11.4	**-0.0364**	0.008	20.8	ug/L	518	Standard
	Sb	123	219.5	27.2	**0.0364**	0.013	37.0	ug/L	127	Standard
	Ba	135	15089.5	2.8	**7.7036**	0.230	3.0	ug/L	45	Standard
	Ce	140	101663.3	0.5				ug/L	298	Standard
>	**Tb**	159	912358.4	0.5				ug/L	977942	Standard
	Ho	165	1823.4	12.2				ug/L	12	Standard
	Tl	203	96.3	39.3	**0.0070**	0.005	65.1	ug/L	31	Standard
	Tl	205	60.0	22.0	**0.0062**	0.002	29.8	ug/L	20	Standard
	Pb	206	5271.9	0.9	**0.9471**	0.013	1.4	ug/L	286	Standard
	Pb	207	4329.0	2.2	**0.8620**	0.023	2.7	ug/L	257	Standard
	Pb	208	18126.6	1.5	**0.9354**	0.019	2.1	ug/L	956	Standard
	U	238	313.3	4.9	**0.0423**	0.002	4.7	ug/L	17	Standard
>	**Bi**	209	537668.4	0.5				ug/L	557080	Standard

Sample ID: L1606078703 WG573333-01
Report Date/Time: Tuesday, June 28, 2016 15:26:41
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	30.0	60.1	**-0.0138**	0.025	183.5	mg/L	32	Standard
K	39	53.3	37.9	**0.0534**	0.067	124.6	mg/L	30	Standard
Ca	43	48.3	6.0	**2.6636**	0.532	20.0	mg/L	48	Standard
Fe	54	632.5	2.4	**0.4873**	0.016	3.2	mg/L	159	Standard
Fe	57	343.3	3.0	**0.6393**	0.010	1.6	mg/L	195	Standard
> **Sc-1**	45	36268.4	3.8				mg/L	39543	Standard
Cl	35	30788.6	1.3				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	1480.1	1.8				ug/L	1647	Standard
P	31	7060.0	1.2				ug/L	15231	Standard
S	34	2285.2	6.0				ug/L	1845	Standard
Sr	88	48.3	21.5				ug/L	70	Standard
C	12	146.7	76.6				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	3101.1	5.7				mg/L	12	Standard
Ho-1	165	1823.4	12.2				mg/L	12	Standard
Er	166	1723.4	3.7				mg/L	20	Standard
I	127	1973.5	19.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		95.369	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.131	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078703 WG573333-01
Report Date/Time: Tuesday, June 28, 2016 15:26:41
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.572
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.515
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078704S WG573333-05
Sample Date/Time: Tuesday, June 28, 2016 15:27:35
Number of Replicates: 3
Autosampler Position: 224
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77117.5	2.8				ug/L	81760	Standard
	Be	9	2866.9	4.2	**3.0501**	0.044	1.4	ug/L	17	Standard
	Al	27	530634.0	2.5	**3.6356**	0.032	0.9	ug/L	1410	Standard
	Sc	45	36769.6	1.1				ug/L	39543	Standard
	Ti	47	373.0	7.6	**1.4053**	0.111	7.9	ug/L	39	Standard
	V	51	27357.6	1.0	**3.9153**	0.028	0.7	ug/L	1415	Standard
	Cr	52	27754.1	2.5	**3.1178**	0.072	2.3	ug/L	6946	Standard
	Cr	53	3205.3	2.9	**3.4822**	0.093	2.7	ug/L	328	Standard
	Mn	55	199454.4	1.6	**32.0700**	0.433	1.4	ug/L	823	Standard
	Co	59	21527.3	1.4	**3.1772**	0.038	1.2	ug/L	252	Standard
	Ni	60	6879.6	1.3	**3.7203**	0.017	0.5	ug/L	192	Standard
	Cu	65	6482.7	2.3	**3.4817**	0.050	1.4	ug/L	141	Standard
	Zn	66	6297.6	0.7	**5.1896**	0.074	1.4	ug/L	491	Standard
>	**Ge**	72	491870.4	0.9				ug/L	515087	Standard
	As	75	2731.5	2.2	**2.5846**	0.063	2.4	ug/L	20	Standard
	Se	82	265.3	2.1	**2.2945**	0.068	3.0	ug/L	15	Standard
	Se-1	77	206.7	1.5	**2.1542**	0.060	2.8	ug/L	48	Standard
>	**Ga**	71	993.4	9.7				mg/L	90	Standard
	Rb	85	7418.5	3.4				ug/L	23	Standard
	Y	89	432036.7	0.9				ug/L	448510	Standard
>	**Rh**	103	6.7	86.6				ug/L	15	Standard
	Mo	98	97.6	12.4	**0.0190**	0.004	23.4	ug/L	11	Standard
	Ag	107	16692.5	1.2	**2.6046**	0.053	2.0	ug/L	134	Standard
	Cd	111	5558.2	2.0	**2.7376**	0.075	2.7	mg/L	7	Standard
	Cd	114	13660.0	0.7	**2.7659**	0.042	1.5	ug/L	16	Standard
>	**In**	115	579914.2	0.8				ug/L	611117	Standard
	Sn	118	355.0	9.2	**-0.0387**	0.005	13.8	ug/L	518	Standard
	Sb	123	2408.5	2.5	**0.5252**	0.017	3.2	ug/L	127	Standard
	Ba	135	18788.0	0.3	**9.5654**	0.079	0.8	ug/L	45	Standard
	Ce	140	97333.3	0.8				ug/L	298	Standard
>	**Tb**	159	916801.1	0.1				ug/L	977942	Standard
	Ho	165	2003.5	4.3				ug/L	12	Standard
	Tl	203	21211.2	1.5	**2.5655**	0.023	0.9	ug/L	31	Standard
	Tl	205	18352.8	1.2	**2.5587**	0.022	0.9	ug/L	20	Standard
	Pb	206	18796.3	1.3	**3.4874**	0.034	1.0	ug/L	286	Standard
	Pb	207	16054.8	0.9	**3.3044**	0.015	0.5	ug/L	257	Standard
	Pb	208	66020.8	0.7	**3.5136**	0.008	0.2	ug/L	956	Standard
	U	238	18577.7	1.3	**2.6444**	0.019	0.7	ug/L	17	Standard
>	**Bi**	209	545605.1	0.6				ug/L	557080	Standard

Sample ID: L1606078704S WG573333-05
Report Date/Time: Tuesday, June 28, 2016 15:29:52
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**23.0608**	19.970	86.6	mg/L	0	Standard
Mg	24	33.3	17.3	**-0.0095**	0.008	83.3	mg/L	32	Standard
K	39	50.0	30.0	**0.0407**	0.057	139.0	mg/L	30	Standard
Ca	43	23.3	86.6	**-1.5565**	3.262	209.6	mg/L	48	Standard
Fe	54	789.3	5.9	**0.6315**	0.054	8.6	mg/L	159	Standard
Fe	57	388.3	19.5	**0.7873**	0.291	37.0	mg/L	195	Standard
> **Sc-1**	45	36769.6	1.1				mg/L	39543	Standard
Cl	35	31616.3	0.3				ug/L	27642	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	1516.7	10.3				ug/L	1647	Standard
P	31	7133.3	2.5				ug/L	15231	Standard
S	34	2288.5	0.7				ug/L	1845	Standard
Sr	88	63.3	12.1				ug/L	70	Standard
C	12	90.0	50.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	3099.8	4.3				mg/L	12	Standard
Ho-1	165	2003.5	4.3				mg/L	12	Standard
Er	166	1750.1	2.1				mg/L	20	Standard
I	127	2031.8	18.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.322	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.493	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078704S WG573333-05
Report Date/Time: Tuesday, June 28, 2016 15:29:52
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.894
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.940
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078705SD WG573333-06
Sample Date/Time: Tuesday, June 28, 2016 15:30:46
Number of Replicates: 3
Autosampler Position: 225
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74501.2	2.9				ug/L	81760	Standard
	Be	9	2816.9	5.3	**3.1033**	0.143	4.6	ug/L	17	Standard
	Al	27	652490.2	1.4	**4.6325**	0.073	1.6	ug/L	1410	Standard
	Sc	45	35373.0	4.0				ug/L	39543	Standard
	Ti	47	288.0	4.5	**1.0503**	0.068	6.5	ug/L	39	Standard
	V	51	27556.1	4.5	**3.9507**	0.115	2.9	ug/L	1415	Standard
	Cr	52	29741.5	2.7	**3.4242**	0.036	1.1	ug/L	6946	Standard
	Cr	53	3408.7	4.0	**3.7308**	0.095	2.5	ug/L	328	Standard
	Mn	55	239390.4	1.7	**38.5872**	0.363	0.9	ug/L	823	Standard
	Co	59	21175.8	2.9	**3.1300**	0.064	2.1	ug/L	252	Standard
	Ni	60	7152.0	2.5	**3.8782**	0.013	0.3	ug/L	192	Standard
	Cu	65	6817.5	3.4	**3.6719**	0.046	1.3	ug/L	141	Standard
	Zn	66	6491.4	2.6	**5.3843**	0.067	1.2	ug/L	491	Standard
>	**Ge**	72	491074.4	2.2				ug/L	515087	Standard
	As	75	2669.0	1.7	**2.5305**	0.018	0.7	ug/L	20	Standard
	Se	82	259.6	0.9	**2.2467**	0.073	3.2	ug/L	15	Standard
	Se-1	77	214.7	9.1	**2.2690**	0.247	10.9	ug/L	48	Standard
>	**Ga**	71	1028.4	8.9				mg/L	90	Standard
	Rb	85	7220.1	2.0				ug/L	23	Standard
	Y	89	434100.4	1.8				ug/L	448510	Standard
>	**Rh**	103	11.7	24.7				ug/L	15	Standard
	Mo	98	220.7	3.6	**0.0685**	0.004	5.3	ug/L	11	Standard
	Ag	107	16241.0	2.2	**2.5784**	0.042	1.6	ug/L	134	Standard
	Cd	111	5563.7	4.0	**2.7880**	0.091	3.3	mg/L	7	Standard
	Cd	114	13468.2	1.0	**2.7750**	0.021	0.8	ug/L	16	Standard
>	**In**	115	569849.5	0.7				ug/L	611117	Standard
	Sn	118	353.3	16.1	**-0.0379**	0.010	26.3	ug/L	518	Standard
	Sb	123	2719.7	5.6	**0.6052**	0.030	5.0	ug/L	127	Standard
	Ba	135	20622.4	3.5	**10.6862**	0.314	2.9	ug/L	45	Standard
	Ce	140	98046.3	2.6				ug/L	298	Standard
>	**Tb**	159	907491.8	2.5				ug/L	977942	Standard
	Ho	165	2030.1	2.1				ug/L	12	Standard
	Tl	203	20792.6	2.1	**2.5412**	0.024	1.0	ug/L	31	Standard
	Tl	205	18564.7	0.9	**2.6157**	0.018	0.7	ug/L	20	Standard
	Pb	206	19134.8	4.2	**3.5891**	0.137	3.8	ug/L	286	Standard
	Pb	207	16318.5	3.9	**3.3949**	0.097	2.9	ug/L	257	Standard
	Pb	208	68373.0	0.9	**3.6800**	0.027	0.7	ug/L	956	Standard
	U	238	18425.5	1.4	**2.6505**	0.039	1.5	ug/L	17	Standard
>	**Bi**	209	539937.0	1.4				ug/L	557080	Standard

Sample ID: L1606078705SD WG573333-06
Report Date/Time: Tuesday, June 28, 2016 15:33:03
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	30.0	44.1	**-0.0120**	0.022	182.1	mg/L	32	Standard
K	39	43.3	29.0	**0.0214**	0.044	203.2	mg/L	30	Standard
Ca	43	31.7	50.8	**-0.0240**	2.524	10500.7	mg/L	48	Standard
Fe	54	837.2	3.3	**0.7115**	0.061	8.6	mg/L	159	Standard
Fe	57	390.0	5.9	**0.8507**	0.142	16.6	mg/L	195	Standard
Sc-1	45	35373.0	4.0				mg/L	39543	Standard
Cl	35	30539.4	1.3				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	1456.7	10.3				ug/L	1647	Standard
P	31	7050.0	3.7				ug/L	15231	Standard
S	34	2141.8	2.3				ug/L	1845	Standard
Sr	88	61.7	12.4				ug/L	70	Standard
C	12	110.0	32.8				mg/L	113	Standard
N	14	6.7	86.6				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	3163.0	6.4				mg/L	12	Standard
Ho-1	165	2030.1	2.1				mg/L	12	Standard
Er	166	1753.4	8.6				mg/L	20	Standard
I	127	1965.1	14.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		91.122	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.338	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606078705SD WG573333-06
Report Date/Time: Tuesday, June 28, 2016 15:33:03
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		93.247
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.923
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078706
Sample Date/Time: Tuesday, June 28, 2016 15:33:58
Number of Replicates: 3
Autosampler Position: 226
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76028.8	3.5				ug/L	81760	Standard
	Be	9	38.3	19.9	**0.0318**	0.009	27.2	ug/L	17	Standard
	Al	27	238262.4	1.5	**1.6489**	0.050	3.0	ug/L	1410	Standard
	Sc	45	37198.9	1.1				ug/L	39543	Standard
	Ti	47	135.7	11.8	**0.3914**	0.065	16.6	ug/L	39	Standard
	V	51	5276.5	1.4	**0.5679**	0.012	2.1	ug/L	1415	Standard
	Cr	52	6498.4	1.5	**-0.1047**	0.015	14.1	ug/L	6946	Standard
	Cr	53	391.7	9.7	**0.1138**	0.045	39.3	ug/L	328	Standard
	Mn	55	119285.7	3.0	**18.6233**	0.543	2.9	ug/L	823	Standard
	Co	59	1493.7	2.3	**0.1855**	0.005	2.9	ug/L	252	Standard
	Ni	60	894.4	1.7	**0.3795**	0.008	2.0	ug/L	192	Standard
	Cu	65	697.3	2.0	**0.2836**	0.007	2.5	ug/L	141	Standard
	Zn	66	2625.6	1.0	**1.6163**	0.028	1.8	ug/L	491	Standard
>	**Ge**	72	504872.7	0.1				ug/L	515087	Standard
	As	75	131.9	5.5	**0.1599**	0.007	4.3	ug/L	20	Standard
	Se	82	16.6	26.9	**0.0277**	0.040	143.8	ug/L	15	Standard
	Se-1	77	42.3	11.2	**-0.1455**	0.065	44.4	ug/L	48	Standard
>	**Ga**	71	331.7	11.3				mg/L	90	Standard
	Rb	85	2800.3	5.6				ug/L	23	Standard
	Y	89	425969.9	0.3				ug/L	448510	Standard
>	**Rh**	103	21.7	96.1				ug/L	15	Standard
	Mo	98	33.2	18.6	**-0.0062**	0.003	40.5	ug/L	11	Standard
	Ag	107	87.7	13.4	**-0.0025**	0.002	77.6	ug/L	134	Standard
	Cd	111	11.3	20.5	**0.0026**	0.001	43.1	mg/L	7	Standard
	Cd	114	30.9	19.2	**0.0013**	0.001	87.3	ug/L	16	Standard
>	**In**	115	583702.0	1.1				ug/L	611117	Standard
	Sn	118	278.3	22.7	**-0.0527**	0.011	20.3	ug/L	518	Standard
	Sb	123	103.7	16.2	**0.0102**	0.004	35.3	ug/L	127	Standard
	Ba	135	9187.1	1.4	**4.6338**	0.016	0.3	ug/L	45	Standard
	Ce	140	45456.4	1.3				ug/L	298	Standard
>	**Tb**	159	930779.6	1.7				ug/L	977942	Standard
	Ho	165	850.0	2.4				ug/L	12	Standard
	Tl	203	23.3	23.6	**-0.0020**	0.001	31.7	ug/L	31	Standard
	Tl	205	35.0	37.8	**0.0025**	0.002	71.7	ug/L	20	Standard
	Pb	206	2536.5	1.2	**0.4062**	0.009	2.2	ug/L	286	Standard
	Pb	207	2050.5	1.7	**0.3630**	0.010	2.7	ug/L	257	Standard
	Pb	208	8421.9	1.7	**0.3865**	0.010	2.6	ug/L	956	Standard
	U	238	103.3	13.1	**0.0115**	0.002	17.1	ug/L	17	Standard
>	**Bi**	209	556617.5	0.8				ug/L	557080	Standard

Sample ID: L1606078706
Report Date/Time: Tuesday, June 28, 2016 15:36:15
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**11.3116**	19.584	173.1	mg/L	0	Standard
Mg	24	31.7	9.1	**-0.0123**	0.005	37.6	mg/L	32	Standard
K	39	50.0	10.0	**0.0383**	0.020	51.6	mg/L	30	Standard
Ca	43	18.3	103.3	**-2.3989**	3.033	126.4	mg/L	48	Standard
Fe	54	317.8	14.7	**0.1684**	0.043	25.3	mg/L	159	Standard
Fe	57	280.0	19.3	**0.3807**	0.196	51.6	mg/L	195	Standard
> **Sc-1**	45	37198.9	1.1				mg/L	39543	Standard
Cl	35	30150.0	0.9				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	1330.1	6.0				ug/L	1647	Standard
P	31	7115.0	3.6				ug/L	15231	Standard
S	34	2203.5	6.4				ug/L	1845	Standard
Sr	88	73.3	23.9				ug/L	70	Standard
C	12	136.7	33.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	1461.3	2.6				mg/L	12	Standard
Ho-1	165	850.0	2.4				mg/L	12	Standard
Er	166	740.0	2.3				mg/L	20	Standard
I	127	1331.7	15.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		92.990	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.017	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078706
Report Date/Time: Tuesday, June 28, 2016 15:36:15
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.514
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.917
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078708
Sample Date/Time: Tuesday, June 28, 2016 15:37:09
Number of Replicates: 3
Autosampler Position: 227
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75812.6	2.8				ug/L	81760	Standard
	Be	9	48.3	21.5	**0.0428**	0.011	26.6	ug/L	17	Standard
	Al	27	1071474.1	1.0	**7.4874**	0.284	3.8	ug/L	1410	Standard
	Sc	45	37513.1	2.1				ug/L	39543	Standard
	Ti	47	158.0	12.8	**0.5014**	0.093	18.5	ug/L	39	Standard
	V	51	6689.1	0.4	**0.8020**	0.010	1.2	ug/L	1415	Standard
	Cr	52	7515.2	0.4	**0.0743**	0.008	10.8	ug/L	6946	Standard
	Cr	53	588.3	8.1	**0.3608**	0.056	15.4	ug/L	328	Standard
	Mn	55	205213.7	2.8	**33.0324**	1.280	3.9	ug/L	823	Standard
	Co	59	2212.5	3.2	**0.2987**	0.013	4.3	ug/L	252	Standard
	Ni	60	1404.4	3.4	**0.6767**	0.035	5.1	ug/L	192	Standard
	Cu	65	1102.0	3.5	**0.5169**	0.018	3.4	ug/L	141	Standard
	Zn	66	2607.9	0.9	**1.6660**	0.041	2.5	ug/L	491	Standard
>	**Ge**	72	491532.3	1.1				ug/L	515087	Standard
	As	75	127.5	11.1	**0.1592**	0.014	8.7	ug/L	20	Standard
	Se	82	17.3	29.3	**0.0385**	0.047	123.2	ug/L	15	Standard
	Se-1	77	43.3	11.6	**-0.1157**	0.075	64.9	ug/L	48	Standard
>	**Ga**	71	450.0	13.5				mg/L	90	Standard
	Rb	85	3788.8	1.3				ug/L	23	Standard
	Y	89	433725.3	2.1				ug/L	448510	Standard
>	**Rh**	103	18.3	31.5				ug/L	15	Standard
	Mo	98	38.5	4.2	**-0.0045**	0.000	10.9	ug/L	11	Standard
	Ag	107	93.3	12.0	**-0.0020**	0.002	78.2	ug/L	134	Standard
	Cd	111	23.9	23.2	**0.0085**	0.003	31.3	mg/L	7	Standard
	Cd	114	45.7	16.0	**0.0041**	0.001	34.3	ug/L	16	Standard
>	**In**	115	598611.3	0.9				ug/L	611117	Standard
	Sn	118	346.7	46.9	**-0.0421**	0.028	66.0	ug/L	518	Standard
	Sb	123	86.1	26.8	**0.0058**	0.005	89.2	ug/L	127	Standard
	Ba	135	10124.4	4.3	**4.9826**	0.248	5.0	ug/L	45	Standard
	Ce	140	56269.4	3.6				ug/L	298	Standard
>	**Tb**	159	935742.2	1.5				ug/L	977942	Standard
	Ho	165	1033.4	8.1				ug/L	12	Standard
	Tl	203	29.7	5.1	**-0.0013**	0.000	18.3	ug/L	31	Standard
	Tl	205	31.7	18.2	**0.0020**	0.001	35.8	ug/L	20	Standard
	Pb	206	3055.6	2.0	**0.5003**	0.022	4.4	ug/L	286	Standard
	Pb	207	2503.2	3.8	**0.4542**	0.030	6.6	ug/L	257	Standard
	Pb	208	10181.4	1.0	**0.4779**	0.016	3.3	ug/L	956	Standard
	U	238	166.3	6.9	**0.0202**	0.002	9.9	ug/L	17	Standard
>	**Bi**	209	558847.5	2.0				ug/L	557080	Standard

Sample ID: L1606078708
Report Date/Time: Tuesday, June 28, 2016 15:39:26
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	26.7	28.6	**-0.0198**	0.010	52.5	mg/L	32	Standard
K	39	46.7	22.3	**0.0253**	0.040	158.3	mg/L	30	Standard
Ca	43	46.7	43.3	**2.1703**	3.421	157.6	mg/L	48	Standard
Fe	54	430.9	12.9	**0.2746**	0.059	21.6	mg/L	159	Standard
Fe	57	336.7	7.3	**0.5728**	0.068	12.0	mg/L	195	Standard
Sc-1	45	37513.1	2.1				mg/L	39543	Standard
Cl	35	31262.2	0.8				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	1390.1	8.3				ug/L	1647	Standard
P	31	7195.0	0.6				ug/L	15231	Standard
S	34	2350.2	4.3				ug/L	1845	Standard
Sr	88	86.7	13.3				ug/L	70	Standard
C	12	130.0	15.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	1824.8	9.6				mg/L	12	Standard
Ho-1	165	1033.4	8.1				mg/L	12	Standard
Er	166	946.7	6.2				mg/L	20	Standard
I	127	1363.4	17.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		92.726	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.427	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078708
Report Date/Time: Tuesday, June 28, 2016 15:39:26
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	97.954	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	100.317	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078713
Sample Date/Time: Tuesday, June 28, 2016 15:40:21
Number of Replicates: 3
Autosampler Position: 228
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75663.5	2.7				ug/L	81760	Standard
	Be	9	81.7	30.2	**0.0793**	0.028	34.9	ug/L	17	Standard
	Al	27	416574.6	1.3	**2.9066**	0.048	1.7	ug/L	1410	Standard
	Sc	45	36378.7	3.4				ug/L	39543	Standard
	Ti	47	126.3	10.5	**0.3773**	0.057	15.2	ug/L	39	Standard
	V	51	8315.3	0.5	**1.0708**	0.007	0.7	ug/L	1415	Standard
	Cr	52	8765.5	1.6	**0.2875**	0.022	7.8	ug/L	6946	Standard
	Cr	53	728.4	7.9	**0.5445**	0.070	12.8	ug/L	328	Standard
	Mn	55	93455.4	1.8	**15.2441**	0.287	1.9	ug/L	823	Standard
	Co	59	3649.1	3.3	**0.5232**	0.019	3.5	ug/L	252	Standard
	Ni	60	1651.1	5.4	**0.8312**	0.052	6.2	ug/L	192	Standard
	Cu	65	1200.0	2.8	**0.5835**	0.019	3.3	ug/L	141	Standard
	Zn	66	2593.9	1.4	**1.6993**	0.037	2.2	ug/L	491	Standard
>	**Ge**	72	482399.4	0.1				ug/L	515087	Standard
	As	75	167.0	21.6	**0.1989**	0.034	17.3	ug/L	20	Standard
	Se	82	15.9	30.3	**0.0282**	0.045	158.8	ug/L	15	Standard
	Se-1	77	52.3	6.7	**0.0228**	0.050	216.8	ug/L	48	Standard
>	**Ga**	71	573.3	13.4				mg/L	90	Standard
	Rb	85	5119.2	1.4				ug/L	23	Standard
	Y	89	439888.1	1.2				ug/L	448510	Standard
>	**Rh**	103	13.3	78.1				ug/L	15	Standard
	Mo	98	43.4	10.8	**-0.0024**	0.002	79.1	ug/L	11	Standard
	Ag	107	106.7	10.0	**0.0002**	0.002	826.8	ug/L	134	Standard
	Cd	111	22.9	46.1	**0.0082**	0.005	63.6	mg/L	7	Standard
	Cd	114	51.2	44.0	**0.0052**	0.004	84.1	ug/L	16	Standard
>	**In**	115	591171.7	0.9				ug/L	611117	Standard
	Sn	118	328.3	8.7	**-0.0445**	0.005	12.1	ug/L	518	Standard
	Sb	123	88.5	32.5	**0.0066**	0.006	98.7	ug/L	127	Standard
	Ba	135	15979.8	0.8	**7.9763**	0.029	0.4	ug/L	45	Standard
	Ce	140	70630.9	0.7				ug/L	298	Standard
>	**Tb**	159	919155.3	2.1				ug/L	977942	Standard
	Ho	165	1308.4	8.9				ug/L	12	Standard
	Tl	203	43.3	33.8	**0.0005**	0.002	355.9	ug/L	31	Standard
	Tl	205	45.0	58.8	**0.0041**	0.004	93.0	ug/L	20	Standard
	Pb	206	4127.6	1.0	**0.7213**	0.011	1.5	ug/L	286	Standard
	Pb	207	3419.4	2.4	**0.6628**	0.016	2.5	ug/L	257	Standard
	Pb	208	14423.4	1.0	**0.7255**	0.013	1.8	ug/L	956	Standard
	U	238	157.3	5.4	**0.0196**	0.001	6.2	ug/L	17	Standard
>	**Bi**	209	542054.4	0.7				ug/L	557080	Standard

Sample ID: L1606078713
Report Date/Time: Tuesday, June 28, 2016 15:42:38
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	38.3	7.5	**-0.0016**	0.005	296.7	mg/L	32	Standard
K	39	31.7	55.5	**-0.0245**	0.065	264.7	mg/L	30	Standard
Ca	43	18.3	56.8	**-2.3425**	1.620	69.1	mg/L	48	Standard
Fe	54	638.8	9.9	**0.4927**	0.076	15.4	mg/L	159	Standard
Fe	57	368.3	10.9	**0.7248**	0.101	14.0	mg/L	195	Standard
> **Sc-1**	45	36378.7	3.4				mg/L	39543	Standard
Cl	35	31640.3	1.2				ug/L	27642	Standard
Kr	83	2.7	57.3				ug/L	2	Standard
Br	81	1476.7	9.6				ug/L	1647	Standard
P	31	7046.6	3.0				ug/L	15231	Standard
S	34	2278.5	11.1				ug/L	1845	Standard
Sr	88	80.0	28.6				ug/L	70	Standard
C	12	116.7	26.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	2168.9	5.7				mg/L	12	Standard
Ho-1	165	1308.4	8.9				mg/L	12	Standard
Er	166	1140.0	7.8				mg/L	20	Standard
I	127	1385.1	16.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		92.543	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.654	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078713
Report Date/Time: Tuesday, June 28, 2016 15:42:38
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.736
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.303
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078715
Sample Date/Time: Tuesday, June 28, 2016 15:43:32
Number of Replicates: 3
Autosampler Position: 229
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72672.4	3.8				ug/L	81760	Standard
	Be	9	136.7	16.5	**0.1453**	0.027	18.6	ug/L	17	Standard
	Al	27	1143496.8	3.3	**8.3341**	0.197	2.4	ug/L	1410	Standard
	Sc	45	36597.5	3.0				ug/L	39543	Standard
	Ti	47	230.0	1.3	**0.8041**	0.017	2.1	ug/L	39	Standard
	V	51	11413.9	1.0	**1.5144**	0.038	2.5	ug/L	1415	Standard
	Cr	52	15036.8	0.3	**1.2067**	0.029	2.4	ug/L	6946	Standard
	Cr	53	1373.4	5.7	**1.2975**	0.078	6.0	ug/L	328	Standard
	Mn	55	293360.4	0.3	**47.2738**	0.683	1.4	ug/L	823	Standard
	Co	59	6852.2	0.9	**0.9907**	0.018	1.8	ug/L	252	Standard
	Ni	60	4182.9	0.4	**2.2225**	0.021	0.9	ug/L	192	Standard
	Cu	65	2749.2	1.6	**1.4253**	0.010	0.7	ug/L	141	Standard
	Zn	66	5934.8	0.5	**4.8465**	0.046	1.0	ug/L	491	Standard
>	**Ge**	72	491565.9	1.3				ug/L	515087	Standard
	As	75	281.2	9.9	**0.3023**	0.024	7.9	ug/L	20	Standard
	Se	82	16.6	41.5	**0.0313**	0.062	197.6	ug/L	15	Standard
	Se-1	77	52.7	11.4	**0.0136**	0.082	602.5	ug/L	48	Standard
>	**Ga**	71	1726.8	3.8				mg/L	90	Standard
	Rb	85	17453.4	2.4				ug/L	23	Standard
	Y	89	433246.4	2.7				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	40.4	18.8	**-0.0027**	0.003	116.5	ug/L	11	Standard
	Ag	107	118.3	7.6	**0.0030**	0.002	54.7	ug/L	134	Standard
	Cd	111	36.9	12.4	**0.0159**	0.002	13.6	mg/L	7	Standard
	Cd	114	100.1	28.9	**0.0161**	0.006	36.4	ug/L	16	Standard
>	**In**	115	559969.6	1.0				ug/L	611117	Standard
	Sn	118	245.0	14.3	**-0.0567**	0.006	11.1	ug/L	518	Standard
	Sb	123	73.5	14.6	**0.0042**	0.003	63.0	ug/L	127	Standard
	Ba	135	26697.5	0.7	**14.0890**	0.209	1.5	ug/L	45	Standard
	Ce	140	126090.9	1.2				ug/L	298	Standard
>	**Tb**	159	888471.0	0.2				ug/L	977942	Standard
	Ho	165	2013.5	3.3				ug/L	12	Standard
	Tl	203	105.3	16.7	**0.0083**	0.002	26.4	ug/L	31	Standard
	Tl	205	98.3	26.1	**0.0118**	0.004	31.7	ug/L	20	Standard
	Pb	206	7239.7	1.0	**1.3394**	0.022	1.6	ug/L	286	Standard
	Pb	207	5907.2	0.7	**1.2110**	0.009	0.7	ug/L	257	Standard
	Pb	208	24779.9	1.0	**1.3155**	0.008	0.6	ug/L	956	Standard
	U	238	285.0	3.3	**0.0387**	0.001	3.3	ug/L	17	Standard
>	**Bi**	209	531928.0	0.6				ug/L	557080	Standard

Sample ID: L1606078715
Report Date/Time: Tuesday, June 28, 2016 15:45:49
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**11.1992**	19.389	173.1	mg/L	0	Standard
Mg	24	35.0	0.0	**-0.0068**	0.002	22.3	mg/L	32	Standard
K	39	70.0	24.7	**0.1145**	0.065	56.7	mg/L	30	Standard
Ca	43	41.7	48.5	**1.5107**	3.417	226.2	mg/L	48	Standard
Fe	54	1498.6	2.3	**1.3298**	0.075	5.6	mg/L	159	Standard
Fe	57	515.0	3.5	**1.2558**	0.109	8.7	mg/L	195	Standard
> **Sc-1**	45	36597.5	3.0				mg/L	39543	Standard
Cl	35	30785.2	1.5				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	1506.7	9.5				ug/L	1647	Standard
P	31	6974.9	2.8				ug/L	15231	Standard
S	34	2301.8	4.8				ug/L	1845	Standard
Sr	88	70.0	28.6				ug/L	70	Standard
C	12	113.3	31.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	3365.6	4.6				mg/L	12	Standard
Ho-1	165	2013.5	3.3				mg/L	12	Standard
Er	166	1770.1	9.2				mg/L	20	Standard
I	127	1690.1	9.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		88.885	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.434	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078715
Report Date/Time: Tuesday, June 28, 2016 15:45:49
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		91.631
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.485
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078716
Sample Date/Time: Tuesday, June 28, 2016 15:46:44
Number of Replicates: 3
Autosampler Position: 230
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72657.2	3.0				ug/L	81760	Standard
	Be	9	110.0	13.6	**0.1151**	0.019	16.6	ug/L	17	Standard
	Al	27	964168.3	0.2	**7.0283**	0.202	2.9	ug/L	1410	Standard
	Sc	45	36550.7	1.7				ug/L	39543	Standard
	Ti	47	122.3	2.9	**0.3511**	0.018	5.2	ug/L	39	Standard
	V	51	5582.1	3.3	**0.6361**	0.029	4.5	ug/L	1415	Standard
	Cr	52	12113.5	1.9	**0.7686**	0.019	2.4	ug/L	6946	Standard
	Cr	53	1075.0	5.2	**0.9442**	0.085	9.0	ug/L	328	Standard
	Mn	55	296391.3	0.6	**47.8289**	1.096	2.3	ug/L	823	Standard
	Co	59	5108.9	2.5	**0.7319**	0.032	4.3	ug/L	252	Standard
	Ni	60	3102.7	1.4	**1.6241**	0.056	3.4	ug/L	192	Standard
	Cu	65	3067.6	0.7	**1.6034**	0.043	2.7	ug/L	141	Standard
	Zn	66	5597.7	1.7	**4.5323**	0.178	3.9	ug/L	491	Standard
>	**Ge**	72	490975.8	1.9				ug/L	515087	Standard
	As	75	34.9	42.7	**0.0731**	0.015	20.1	ug/L	20	Standard
	Se	82	17.1	13.6	**0.0368**	0.020	54.2	ug/L	15	Standard
	Se-1	77	53.7	6.0	**0.0281**	0.032	114.8	ug/L	48	Standard
>	**Ga**	71	1235.1	6.5				mg/L	90	Standard
	Rb	85	7633.6	0.9				ug/L	23	Standard
	Y	89	431776.8	2.6				ug/L	448510	Standard
>	**Rh**	103	15.0	66.7				ug/L	15	Standard
	Mo	98	19.3	30.7	**-0.0115**	0.002	20.4	ug/L	11	Standard
	Ag	107	114.7	7.5	**0.0018**	0.001	48.3	ug/L	134	Standard
	Cd	111	65.3	10.2	**0.0294**	0.003	8.8	mg/L	7	Standard
	Cd	114	154.6	13.1	**0.0266**	0.003	12.9	ug/L	16	Standard
>	**In**	115	577281.8	2.6				ug/L	611117	Standard
	Sn	118	261.7	10.5	**-0.0551**	0.004	6.8	ug/L	518	Standard
	Sb	123	70.5	12.0	**0.0029**	0.001	50.3	ug/L	127	Standard
	Ba	135	31515.4	1.1	**16.1422**	0.456	2.8	ug/L	45	Standard
	Ce	140	108483.7	1.5				ug/L	298	Standard
>	**Tb**	159	919053.3	0.1				ug/L	977942	Standard
	Ho	165	2146.8	3.5				ug/L	12	Standard
	Tl	203	55.7	15.3	**0.0020**	0.001	51.1	ug/L	31	Standard
	Tl	205	56.7	10.2	**0.0057**	0.001	13.8	ug/L	20	Standard
	Pb	206	8383.7	0.8	**1.5358**	0.026	1.7	ug/L	286	Standard
	Pb	207	6950.9	2.1	**1.4122**	0.046	3.3	ug/L	257	Standard
	Pb	208	28623.6	0.6	**1.5044**	0.028	1.9	ug/L	956	Standard
	U	238	278.7	6.7	**0.0371**	0.002	6.0	ug/L	17	Standard
>	**Bi**	209	540360.9	1.3				ug/L	557080	Standard

Sample ID: L1606078716
Report Date/Time: Tuesday, June 28, 2016 15:49:01
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**23.0232**	19.940	86.6	mg/L	0	Standard
Mg	24	31.7	36.5	**-0.0117**	0.016	134.2	mg/L	32	Standard
K	39	40.0	57.3	**0.0042**	0.082	1950.6	mg/L	30	Standard
Ca	43	45.0	19.2	**2.0338**	1.285	63.2	mg/L	48	Standard
Fe	54	558.7	3.5	**0.4103**	0.028	6.8	mg/L	159	Standard
Fe	57	366.7	19.3	**0.7123**	0.240	33.7	mg/L	195	Standard
Sc-1	45	36550.7	1.7				mg/L	39543	Standard
Cl	35	30424.5	0.4				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	1540.1	8.3				ug/L	1647	Standard
P	31	6878.2	1.3				ug/L	15231	Standard
S	34	2278.5	4.6				ug/L	1845	Standard
Sr	88	71.7	29.0				ug/L	70	Standard
C	12	150.0	33.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	0	Standard
Dy	164	3587.7	5.2				mg/L	12	Standard
Ho-1	165	2146.8	3.5				mg/L	12	Standard
Er	166	1936.8	10.5				mg/L	20	Standard
I	127	1943.5	8.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		88.866	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.319	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606078716
Report Date/Time: Tuesday, June 28, 2016 15:49:01
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.463
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.999
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606078717
Sample Date/Time: Tuesday, June 28, 2016 15:49:55
Number of Replicates: 3
Autosampler Position: 231
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72923.4	1.2				ug/L	81760	Standard
	Be	9	50.0	43.6	**0.0466**	0.024	52.0	ug/L	17	Standard
	Al	27	245816.4	0.3	**1.7738**	0.027	1.5	ug/L	1410	Standard
	Sc	45	37411.1	1.3				ug/L	39543	Standard
	Ti	47	150.7	7.6	**0.4619**	0.045	9.8	ug/L	39	Standard
	V	51	5992.6	2.2	**0.6857**	0.015	2.1	ug/L	1415	Standard
	Cr	52	7687.3	1.9	**0.0860**	0.027	31.9	ug/L	6946	Standard
	Cr	53	518.3	10.1	**0.2697**	0.061	22.7	ug/L	328	Standard
	Mn	55	119301.5	0.3	**18.8992**	0.085	0.4	ug/L	823	Standard
	Co	59	2188.2	2.9	**0.2910**	0.008	2.8	ug/L	252	Standard
	Ni	60	1191.7	0.5	**0.5500**	0.007	1.2	ug/L	192	Standard
	Cu	65	843.0	2.1	**0.3684**	0.010	2.8	ug/L	141	Standard
	Zn	66	2526.9	1.6	**1.5588**	0.051	3.2	ug/L	491	Standard
>	**Ge**	72	497647.3	0.6				ug/L	515087	Standard
	As	75	96.2	13.8	**0.1289**	0.013	9.7	ug/L	20	Standard
	Se	82	10.6	36.2	**-0.0241**	0.035	144.8	ug/L	15	Standard
	Se-1	77	46.0	14.3	**-0.0871**	0.088	100.8	ug/L	48	Standard
>	**Ga**	71	388.3	9.7				mg/L	90	Standard
	Rb	85	3552.1	4.5				ug/L	23	Standard
	Y	89	408184.3	3.4				ug/L	448510	Standard
>	**Rh**	103	16.7	17.3				ug/L	15	Standard
	Mo	98	62.0	6.5	**0.0053**	0.001	22.5	ug/L	11	Standard
	Ag	107	95.7	11.7	**-0.0010**	0.002	195.1	ug/L	134	Standard
	Cd	111	10.9	9.2	**0.0025**	0.001	23.6	mg/L	7	Standard
	Cd	114	23.2	24.4	**-0.0002**	0.001	811.4	ug/L	16	Standard
>	**In**	115	574185.6	1.8				ug/L	611117	Standard
	Sn	118	255.0	18.7	**-0.0561**	0.008	14.6	ug/L	518	Standard
	Sb	123	68.2	5.6	**0.0025**	0.001	35.8	ug/L	127	Standard
	Ba	135	10138.1	0.2	**5.2024**	0.085	1.6	ug/L	45	Standard
	Ce	140	49457.3	0.9				ug/L	298	Standard
>	**Tb**	159	906877.5	2.2				ug/L	977942	Standard
	Ho	165	853.4	7.8				ug/L	12	Standard
	Tl	203	26.0	16.8	**-0.0016**	0.001	32.4	ug/L	31	Standard
	Tl	205	21.7	26.6	**0.0008**	0.001	106.7	ug/L	20	Standard
	Pb	206	2799.3	1.6	**0.4682**	0.011	2.3	ug/L	286	Standard
	Pb	207	2286.8	2.0	**0.4236**	0.011	2.7	ug/L	257	Standard
	Pb	208	9380.8	1.4	**0.4502**	0.009	2.0	ug/L	956	Standard
	U	238	106.7	2.9	**0.0123**	0.000	3.6	ug/L	17	Standard
>	**Bi**	209	542636.2	0.5				ug/L	557080	Standard

Sample ID: L1606078717
Report Date/Time: Tuesday, June 28, 2016 15:52:12
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**22.4620**	19.449	86.6	mg/L	0	Standard
Mg	24	26.7	21.7	**-0.0196**	0.008	42.2	mg/L	32	Standard
K	39	36.7	7.9	**-0.0105**	0.010	94.8	mg/L	30	Standard
Ca	43	36.7	56.8	**0.5265**	3.309	628.5	mg/L	48	Standard
Fe	54	389.3	2.1	**0.2354**	0.013	5.4	mg/L	159	Standard
Fe	57	280.0	3.1	**0.3744**	0.024	6.3	mg/L	195	Standard
> **Sc-1**	45	37411.1	1.3				mg/L	39543	Standard
Cl	35	30529.4	0.4				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	1530.1	5.4				ug/L	1647	Standard
P	31	7263.4	2.6				ug/L	15231	Standard
S	34	2140.2	5.9				ug/L	1845	Standard
Sr	88	53.3	23.6				ug/L	70	Standard
C	12	100.0	75.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	1448.1	5.9				mg/L	12	Standard
Ho-1	165	853.4	7.8				mg/L	12	Standard
Er	166	736.7	15.9				mg/L	20	Standard
I	127	6059.5	4.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		89.192	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.614	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606078717
Report Date/Time: Tuesday, June 28, 2016 15:52:12
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		93.957
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.407
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 15:53:08
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80037.6	2.2				ug/L	81760	Standard
	Be	9	50415.4	0.4	**51.8632**	1.062	2.0	ug/L	17	Standard
	Al	27	7611642.9	1.2	**50.4371**	0.677	1.3	ug/L	1410	Standard
	Sc	45	41097.3	0.2				ug/L	39543	Standard
	Ti	47	25369.9	1.7	**101.8781**	2.068	2.0	ug/L	39	Standard
	V	51	351293.9	2.1	**50.3394**	0.543	1.1	ug/L	1415	Standard
	Cr	52	361193.1	2.4	**50.8341**	0.636	1.3	ug/L	6946	Standard
	Cr	53	45189.0	2.1	**51.1421**	0.625	1.2	ug/L	328	Standard
	Mn	55	322540.4	1.4	**49.6098**	0.122	0.2	ug/L	823	Standard
	Co	59	356756.8	0.9	**50.7536**	0.285	0.6	ug/L	252	Standard
	Ni	60	97064.3	2.1	**51.4369**	0.357	0.7	ug/L	192	Standard
	Cu	65	97360.2	2.2	**51.1525**	0.591	1.2	ug/L	141	Standard
	Zn	66	56918.1	2.1	**51.1056**	0.446	0.9	ug/L	491	Standard
>	**Ge**	72	515026.5	1.4				ug/L	515087	Standard
	As	75	56903.6	1.1	**50.6598**	0.179	0.4	ug/L	20	Standard
	Se	82	5856.2	1.8	**50.7515**	0.222	0.4	ug/L	15	Standard
	Se-1	77	3927.8	0.7	**51.4293**	0.511	1.0	ug/L	48	Standard
>	**Ga**	71	66.7	30.3				mg/L	90	Standard
	Rb	85	760.0	4.6				ug/L	23	Standard
	Y	89	447852.9	1.0				ug/L	448510	Standard
>	**Rh**	103	25.0	34.6				ug/L	15	Standard
	Mo	98	271272.3	0.9	**98.6404**	1.710	1.7	ug/L	11	Standard
	Ag	107	339950.6	2.5	**49.7752**	1.742	3.5	ug/L	134	Standard
	Cd	111	107240.7	1.0	**49.3177**	0.985	2.0	mg/L	7	Standard
	Cd	114	262943.6	0.6	**49.7461**	0.816	1.6	ug/L	16	Standard
>	**In**	115	621725.8	1.1				ug/L	611117	Standard
	Sn	118	299388.1	2.0	**49.6643**	0.590	1.2	ug/L	518	Standard
	Sb	123	239448.5	1.8	**49.8807**	1.390	2.8	ug/L	127	Standard
	Ba	135	102748.9	0.7	**48.8974**	0.441	0.9	ug/L	45	Standard
	Ce	140	90.0	5.6				ug/L	298	Standard
>	**Tb**	159	985013.6	1.3				ug/L	977942	Standard
	Ho	165	16.7	45.8				ug/L	12	Standard
	Tl	203	430595.1	1.2	**49.9408**	0.227	0.5	ug/L	31	Standard
	Tl	205	380549.9	0.8	**50.8302**	0.446	0.9	ug/L	20	Standard
	Pb	206	272066.9	1.5	**49.1343**	0.429	0.9	ug/L	286	Standard
	Pb	207	246215.9	1.8	**49.3013**	0.781	1.6	ug/L	257	Standard
	Pb	208	985827.4	0.9	**51.0241**	0.323	0.6	ug/L	956	Standard
	U	238	363383.0	0.6	**49.5660**	0.508	1.0	ug/L	17	Standard
>	**Bi**	209	570034.9	1.5				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:55:25
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.2894**	17.813	173.1	mg/L	0	Standard
Mg	24	3815.5	2.6	**4.8299**	0.134	2.8	mg/L	32	Standard
K	39	1281.7	1.9	**4.0254**	0.078	1.9	mg/L	30	Standard
Ca	43	63.3	19.9	**3.9099**	1.840	47.1	mg/L	48	Standard
Fe	54	5945.2	1.8	**5.0400**	0.102	2.0	mg/L	159	Standard
Fe	57	1733.4	4.2	**5.0061**	0.238	4.8	mg/L	195	Standard
> **Sc-1**	45	41097.3	0.2				mg/L	39543	Standard
Cl	35	35020.4	1.1				ug/L	27642	Standard
Kr	83	3.0	57.7				ug/L	2	Standard
Br	81	1666.8	6.4				ug/L	1647	Standard
P	31	18748.3	2.5				ug/L	15231	Standard
S	34	2583.6	5.3				ug/L	1845	Standard
Sr	88	53.3	28.6				ug/L	70	Standard
C	12	110.0	24.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.4	101.7				mg/L	12	Standard
Ho-1	165	16.7	45.8				mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	1388.4	22.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	103.726		
	Al	27	100.874		
	Sc	45			
	Ti	47	101.878		
	V	51	100.679		
	Cr	52	101.668		
	Cr	53			
	Mn	55	99.220		
	Co	59	101.507		
	Ni	60	102.874		
	Cu	65	102.305		
	Zn	66	102.211		
>	Ge	72		99.988	
	As	75	101.320		
	Se	82	101.503		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:55:25
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	98.640		
Ag	107	99.550		
Cd	111	98.635		
Cd	114			
In	115			101.736
Sn	118	99.329		
Sb	123	99.761		
Ba	135	97.795		
Ce	140			
Tb	159			
Ho	165			
Tl	203	99.882		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.048		
U	238	99.132		
Bi	209			102.325
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 15:55:25
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 15:56:20
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79843.2	1.6				ug/L	81760	Standard
	Be	9	23.3	75.3	**0.0144**	0.018	125.7	ug/L	17	Standard
	Al	27	2505.2	18.4	**0.0022**	0.003	136.7	ug/L	1410	Standard
	Sc	45	39966.0	1.7				ug/L	39543	Standard
	Ti	47	35.3	11.4	**-0.0209**	0.017	83.6	ug/L	39	Standard
	V	51	1133.0	9.8	**-0.0412**	0.016	39.7	ug/L	1415	Standard
	Cr	52	6206.9	2.0	**-0.1524**	0.017	11.1	ug/L	6946	Standard
	Cr	53	315.0	2.7	**0.0227**	0.013	56.9	ug/L	328	Standard
	Mn	55	686.0	7.5	**-0.0412**	0.007	17.6	ug/L	823	Standard
	Co	59	254.3	2.2	**0.0054**	0.001	12.6	ug/L	252	Standard
	Ni	60	171.3	1.2	**-0.0127**	0.001	8.3	ug/L	192	Standard
	Cu	65	123.3	12.6	**-0.0250**	0.008	31.0	ug/L	141	Standard
	Zn	66	449.3	3.1	**-0.4117**	0.013	3.2	ug/L	491	Standard
>	**Ge**	72	507665.8	0.8				ug/L	515087	Standard
	As	75	-25.2	101.2	**0.0174**	0.023	132.3	ug/L	20	Standard
	Se	82	17.2	13.7	**0.0322**	0.020	60.6	ug/L	15	Standard
	Se-1	77	51.3	10.7	**-0.0277**	0.072	260.3	ug/L	48	Standard
>	**Ga**	71	48.3	15.8				mg/L	90	Standard
	Rb	85	30.0	44.1				ug/L	23	Standard
	Y	89	444982.9	1.9				ug/L	448510	Standard
>	**Rh**	103	16.7	34.6				ug/L	15	Standard
	Mo	98	252.7	24.3	**0.0725**	0.022	30.7	ug/L	11	Standard
	Ag	107	129.7	3.6	**0.0027**	0.001	26.2	ug/L	134	Standard
	Cd	111	17.2	9.0	**0.0050**	0.001	14.0	mg/L	7	Standard
	Cd	114	47.7	10.5	**0.0041**	0.001	23.1	ug/L	16	Standard
>	**In**	115	623883.8	0.1				ug/L	611117	Standard
	Sn	118	770.0	14.1	**0.0256**	0.018	70.0	ug/L	518	Standard
	Sb	123	507.4	30.1	**0.0925**	0.032	34.2	ug/L	127	Standard
	Ba	135	38.7	5.4	**-0.0068**	0.001	14.6	ug/L	45	Standard
	Ce	140	40.0	21.7				ug/L	298	Standard
>	**Tb**	159	974680.3	0.5				ug/L	977942	Standard
	Ho	165	13.3	78.1				ug/L	12	Standard
	Tl	203	68.0	6.7	**0.0032**	0.001	15.5	ug/L	31	Standard
	Tl	205	66.7	11.5	**0.0068**	0.001	14.7	ug/L	20	Standard
	Pb	206	289.3	1.4	**-0.0104**	0.001	6.9	ug/L	286	Standard
	Pb	207	250.7	7.8	**-0.0070**	0.004	55.4	ug/L	257	Standard
	Pb	208	1099.7	2.8	**-0.0025**	0.002	60.9	ug/L	956	Standard
	U	238	62.3	42.5	**0.0057**	0.004	64.1	ug/L	17	Standard
>	**Bi**	209	560756.1	0.5				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:58:37
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	33.3	45.8	**-0.0130**	0.021	161.7	mg/L	32	Standard
K	39	33.3	17.3	**-0.0301**	0.018	60.0	mg/L	30	Standard
Ca	43	38.3	27.2	**0.4086**	1.513	370.3	mg/L	48	Standard
Fe	54	141.0	17.0	**-0.0107**	0.022	208.8	mg/L	159	Standard
Fe	57	213.3	13.7	**0.0887**	0.105	118.5	mg/L	195	Standard
> **Sc-1**	45	39966.0	1.7				mg/L	39543	Standard
Cl	35	34581.4	2.4				ug/L	27642	Standard
Kr	83	3.0	66.7				ug/L	2	Standard
Br	81	1706.8	4.2				ug/L	1647	Standard
P	31	18139.2	2.9				ug/L	15231	Standard
S	34	2418.5	8.0				ug/L	1845	Standard
Sr	88	68.3	11.2				ug/L	70	Standard
C	12	130.0	53.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.8	101.7				mg/L	12	Standard
Ho-1	165	13.3	78.1				mg/L	12	Standard
Er	166	3.3	173.2				mg/L	20	Standard
I	127	1121.7	18.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.559	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 15:58:37
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.089
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.660
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: PBW 45 WG573751-04
Sample Date/Time: Tuesday, June 28, 2016 15:59:33
Number of Replicates: 3
Autosampler Position: 232
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80427.9	0.7				ug/L	81760	Standard
	Be	9	20.0	43.3	**0.0107**	0.009	82.0	ug/L	17	Standard
	Al	27	16896.1	3.6	**0.0970**	0.004	4.0	ug/L	1410	Standard
	Sc	45	39630.1	2.3				ug/L	39543	Standard
	Ti	47	34.3	28.0	**-0.0290**	0.037	127.4	ug/L	39	Standard
	V	51	1241.7	7.6	**-0.0303**	0.012	39.4	ug/L	1415	Standard
	Cr	52	8808.2	2.5	**0.1923**	0.015	7.8	ug/L	6946	Standard
	Cr	53	705.0	11.9	**0.4526**	0.107	23.5	ug/L	328	Standard
	Mn	55	5471.7	0.3	**0.6851**	0.014	2.1	ug/L	823	Standard
	Co	59	272.7	7.3	**0.0070**	0.003	48.0	ug/L	252	Standard
	Ni	60	321.0	11.4	**0.0633**	0.019	30.2	ug/L	192	Standard
	Cu	65	232.7	11.6	**0.0301**	0.014	46.8	ug/L	141	Standard
	Zn	66	2018.1	1.4	**0.9905**	0.004	0.4	ug/L	491	Standard
>	**Ge**	72	521638.1	1.4				ug/L	515087	Standard
	As	75	-54.9	36.7	**-0.0079**	0.017	218.9	ug/L	20	Standard
	Se	82	10.0	20.3	**-0.0338**	0.018	54.1	ug/L	15	Standard
	Se-1	77	48.0	14.4	**-0.0888**	0.100	112.7	ug/L	48	Standard
>	**Ga**	71	53.3	28.6				mg/L	90	Standard
	Rb	85	156.7	21.7				ug/L	23	Standard
	Y	89	436462.5	0.8				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	82.2	21.1	**0.0108**	0.006	59.5	ug/L	11	Standard
	Ag	107	93.7	16.2	**-0.0025**	0.002	87.2	ug/L	134	Standard
	Cd	111	11.9	25.2	**0.0025**	0.001	54.5	mg/L	7	Standard
	Cd	114	46.6	47.9	**0.0039**	0.004	108.7	ug/L	16	Standard
>	**In**	115	622987.2	0.5				ug/L	611117	Standard
	Sn	118	873.4	7.7	**0.0429**	0.010	24.2	ug/L	518	Standard
	Sb	123	225.0	36.0	**0.0339**	0.017	49.6	ug/L	127	Standard
	Ba	135	668.7	0.9	**0.2926**	0.004	1.5	ug/L	45	Standard
	Ce	140	103.3	7.4				ug/L	298	Standard
>	**Tb**	159	980754.1	0.6				ug/L	977942	Standard
	Ho	165	13.3	78.1				ug/L	12	Standard
	Tl	203	31.7	30.7	**-0.0012**	0.001	94.5	ug/L	31	Standard
	Tl	205	30.0	86.6	**0.0017**	0.003	203.8	ug/L	20	Standard
	Pb	206	380.3	5.4	**0.0044**	0.003	72.4	ug/L	286	Standard
	Pb	207	331.3	8.5	**0.0076**	0.005	67.3	ug/L	257	Standard
	Pb	208	1302.7	2.5	**0.0063**	0.001	19.3	ug/L	956	Standard
	U	238	25.3	53.6	**0.0004**	0.002	408.2	ug/L	17	Standard
>	**Bi**	209	576453.2	0.7				ug/L	557080	Standard

Sample ID: PBW 45 WG573751-04
Report Date/Time: Tuesday, June 28, 2016 16:01:50
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	33.3	8.7	**-0.0129**	0.003	22.5	mg/L	32	Standard
K	39	48.3	15.8	**0.0218**	0.028	129.2	mg/L	30	Standard
Ca	43	18.3	78.7	**-2.5918**	2.118	81.7	mg/L	48	Standard
Fe	54	122.7	28.9	**-0.0266**	0.030	114.3	mg/L	159	Standard
Fe	57	223.3	24.1	**0.1269**	0.176	138.7	mg/L	195	Standard
Sc-1	45	39630.1	2.3				mg/L	39543	Standard
Cl	35	33807.0	1.7				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	1633.4	2.0				ug/L	1647	Standard
P	31	17535.1	2.2				ug/L	15231	Standard
S	34	2338.5	7.5				ug/L	1845	Standard
Sr	88	75.0	30.6				ug/L	70	Standard
C	12	106.7	19.5				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	9.4	3.0				mg/L	12	Standard
Ho-1	165	13.3	78.1				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	955.0	19.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		98.371	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		101.272	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: PBW 45 WG573751-04
Report Date/Time: Tuesday, June 28, 2016 16:01:50
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.942
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.478
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: LCSW 45 WG573451-05
Sample Date/Time: Tuesday, June 28, 2016 16:02:44
Number of Replicates: 3
Autosampler Position: 233
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81285.8	1.4				ug/L	81760	Standard
	Be	9	49444.0	4.0	**50.0648**	1.744	3.5	ug/L	17	Standard
	Al	27	14440.2	1.6	**0.0798**	0.002	2.1	ug/L	1410	Standard
	Sc	45	40522.5	4.2				ug/L	39543	Standard
	Ti	47	35.7	14.1	**-0.0217**	0.021	96.5	ug/L	39	Standard
	V	51	343445.8	1.5	**49.2175**	1.714	3.5	ug/L	1415	Standard
	Cr	52	355835.3	1.6	**50.0593**	1.026	2.0	ug/L	6946	Standard
	Cr	53	43403.8	4.1	**49.0831**	1.336	2.7	ug/L	328	Standard
	Mn	55	327263.1	1.9	**50.3439**	1.847	3.7	ug/L	823	Standard
	Co	59	357909.9	1.9	**50.9146**	1.589	3.1	ug/L	252	Standard
	Ni	60	94484.4	2.0	**50.0595**	0.992	2.0	ug/L	192	Standard
	Cu	65	95887.5	1.7	**50.3702**	0.997	2.0	ug/L	141	Standard
	Zn	66	56697.9	0.9	**50.8990**	0.801	1.6	ug/L	491	Standard
>	**Ge**	72	515232.3	2.4				ug/L	515087	Standard
	As	75	56137.8	1.3	**49.9742**	1.270	2.5	ug/L	20	Standard
	Se	82	5754.2	0.4	**49.8676**	1.316	2.6	ug/L	15	Standard
	Se-1	77	3847.2	2.4	**50.3547**	1.829	3.6	ug/L	48	Standard
>	**Ga**	71	80.0	38.0				mg/L	90	Standard
	Rb	85	45.0	58.8				ug/L	23	Standard
	Y	89	449389.0	2.2				ug/L	448510	Standard
>	**Rh**	103	28.3	66.8				ug/L	15	Standard
	Mo	98	72.0	12.6	**0.0070**	0.003	48.7	ug/L	11	Standard
	Ag	107	332360.6	2.0	**48.5009**	1.256	2.6	ug/L	134	Standard
	Cd	111	105950.1	2.1	**48.5677**	1.420	2.9	mg/L	7	Standard
	Cd	114	259244.9	2.6	**48.8950**	1.818	3.7	ug/L	16	Standard
>	**In**	115	623755.7	1.2				ug/L	611117	Standard
	Sn	118	950.0	6.3	**0.0554**	0.008	14.9	ug/L	518	Standard
	Sb	123	235787.4	1.4	**48.9583**	1.261	2.6	ug/L	127	Standard
	Ba	135	101234.2	2.5	**48.0272**	1.604	3.3	ug/L	45	Standard
	Ce	140	53.3	5.4				ug/L	298	Standard
>	**Tb**	159	962831.7	2.3				ug/L	977942	Standard
	Ho	165	6.7	114.6				ug/L	12	Standard
	Tl	203	424492.3	2.2	**49.7361**	1.545	3.1	ug/L	31	Standard
	Tl	205	378457.4	1.6	**51.0672**	1.520	3.0	ug/L	20	Standard
	Pb	206	276584.3	1.3	**50.4636**	1.380	2.7	ug/L	286	Standard
	Pb	207	241372.6	1.3	**48.8258**	1.366	2.8	ug/L	257	Standard
	Pb	208	990155.6	1.9	**51.7752**	1.712	3.3	ug/L	956	Standard
	U	238	358184.2	2.6	**49.3578**	1.903	3.9	ug/L	17	Standard
>	**Bi**	209	564381.2	1.4				ug/L	557080	Standard

Sample ID: LCSW 45 WG573451-05
Report Date/Time: Tuesday, June 28, 2016 16:05:01
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.8762**	18.829	173.1	mg/L	0	Standard
Mg	24	28.3	71.3	**-0.0199**	0.028	139.9	mg/L	32	Standard
K	39	45.0	40.1	**0.0061**	0.057	941.7	mg/L	30	Standard
Ca	43	25.0	60.0	**-1.6629**	2.211	133.0	mg/L	48	Standard
Fe	54	141.6	12.0	**-0.0123**	0.010	79.8	mg/L	159	Standard
Fe	57	180.0	5.6	**-0.0319**	0.009	26.7	mg/L	195	Standard
Sc-1	45	40522.5	4.2				mg/L	39543	Standard
Cl	35	33285.9	3.1				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	5364.3	7.6				ug/L	1647	Standard
P	31	17471.7	4.0				ug/L	15231	Standard
S	34	2361.9	6.4				ug/L	1845	Standard
Sr	88	63.3	24.1				ug/L	70	Standard
C	12	126.7	43.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	15.9	34.8				mg/L	12	Standard
Ho-1	165	6.7	114.6				mg/L	12	Standard
Er	166	16.7	69.3				mg/L	20	Standard
I	127	856.7	31.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		99.420	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.028	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 45 WG573451-05
Report Date/Time: Tuesday, June 28, 2016 16:05:01
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.068
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.311
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 45 WG573451-05
Report Date/Time: Tuesday, June 28, 2016 16:05:01
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115101 WG573751-02
Sample Date/Time: Tuesday, June 28, 2016 16:05:56
Number of Replicates: 3
Autosampler Position: 234
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77671.9	1.4				ug/L	81760	Standard
	Be	9	26.7	39.0	**0.0185**	0.011	58.4	ug/L	17	Standard
	Al	27	13955917.4	1.2	**95.3134**	2.493	2.6	ug/L	1410	Standard
	Sc	45	39321.0	0.9				ug/L	39543	Standard
	Ti	47	301.7	10.7	**1.0707**	0.148	13.8	ug/L	39	Standard
	V	51	16198.5	2.2	**2.1650**	0.025	1.2	ug/L	1415	Standard
	Cr	52	10911.9	4.6	**0.5376**	0.044	8.1	ug/L	6946	Standard
	Cr	53	3267.0	16.4	**3.4384**	0.551	16.0	ug/L	328	Standard
	Mn	55	181411.8	1.7	**28.3397**	0.987	3.5	ug/L	823	Standard
	Co	59	976.7	3.3	**0.1102**	0.006	5.4	ug/L	252	Standard
	Ni	60	1471.7	0.7	**0.6905**	0.022	3.2	ug/L	192	Standard
	Cu	65	1872.1	2.6	**0.9119**	0.023	2.5	ug/L	141	Standard
	Zn	66	6550.4	2.0	**5.2541**	0.063	1.2	ug/L	491	Standard
>	**Ge**	72	506178.8	2.0				ug/L	515087	Standard
	As	75	931.2	4.4	**0.8837**	0.051	5.8	ug/L	20	Standard
	Se	82	231.2	4.1	**1.9256**	0.108	5.6	ug/L	15	Standard
	Se-1	77	230.7	3.7	**2.3974**	0.119	5.0	ug/L	48	Standard
>	**Ga**	71	195.0	5.1				mg/L	90	Standard
	Rb	85	19537.6	1.2				ug/L	23	Standard
	Y	89	432314.6	1.1				ug/L	448510	Standard
>	**Rh**	103	61.7	12.4				ug/L	15	Standard
	Mo	98	3291.8	1.3	**1.2330**	0.013	1.0	ug/L	11	Standard
	Ag	107	150.7	8.2	**0.0069**	0.002	29.9	ug/L	134	Standard
	Cd	111	62.2	4.0	**0.0270**	0.001	5.4	mg/L	7	Standard
	Cd	114	153.0	45.9	**0.0254**	0.014	55.2	ug/L	16	Standard
>	**In**	115	594394.0	1.1				ug/L	611117	Standard
	Sn	118	883.4	9.5	**0.0516**	0.013	25.0	ug/L	518	Standard
	Sb	123	639.6	8.2	**0.1265**	0.011	8.6	ug/L	127	Standard
	Ba	135	68518.2	0.3	**34.1005**	0.460	1.3	ug/L	45	Standard
	Ce	140	2645.2	17.9				ug/L	298	Standard
>	**Tb**	159	950271.0	0.7				ug/L	977942	Standard
	Ho	165	53.3	10.8				ug/L	12	Standard
	Tl	203	411.7	14.6	**0.0459**	0.008	17.1	ug/L	31	Standard
	Tl	205	376.7	19.3	**0.0511**	0.011	20.8	ug/L	20	Standard
	Pb	206	1679.8	1.4	**0.2582**	0.003	1.3	ug/L	286	Standard
	Pb	207	1366.4	3.8	**0.2322**	0.009	3.7	ug/L	257	Standard
	Pb	208	5841.1	1.9	**0.2602**	0.004	1.4	ug/L	956	Standard
	U	238	9731.1	2.1	**1.4034**	0.022	1.5	ug/L	17	Standard
>	**Bi**	209	537950.1	0.8				ug/L	557080	Standard

Sample ID: L1606115101 WG573751-02
Report Date/Time: Tuesday, June 28, 2016 16:08:13
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	34736.4	1.4	**46.4420**	0.278	0.6	mg/L	32	Standard
K	39	261.7	6.7	**0.7474**	0.052	7.0	mg/L	30	Standard
Ca	43	178.3	24.5	**21.8549**	6.453	29.5	mg/L	48	Standard
Fe	54	262.3	8.0	**0.1018**	0.020	19.8	mg/L	159	Standard
Fe	57	341.7	2.2	**0.5353**	0.023	4.3	mg/L	195	Standard
> **Sc-1**	45	39321.0	0.9				mg/L	39543	Standard
Cl	35	35275.1	3.8				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	35124.1	4.7				ug/L	1647	Standard
P	31	18164.2	1.2				ug/L	15231	Standard
S	34	2260.2	2.8				ug/L	1845	Standard
Sr	88	231.7	13.9				ug/L	70	Standard
C	12	256.7	41.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	63.6	33.9				mg/L	12	Standard
Ho-1	165	53.3	10.8				mg/L	12	Standard
Er	166	63.3	36.5				mg/L	20	Standard
I	127	311901.7	7.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		95.000	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.271	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115101 WG573751-02
Report Date/Time: Tuesday, June 28, 2016 16:08:13
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.264
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.566
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115103S WG573751-07
Sample Date/Time: Tuesday, June 28, 2016 16:09:07
Number of Replicates: 3
Autosampler Position: 235
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82738.6	0.2				ug/L	81760	Standard
	Be	9	50932.1	1.3	**50.6700**	0.765	1.5	ug/L	17	Standard
	Al	27	14256616.7	2.2	**91.3792**	1.822	2.0	ug/L	1410	Standard
	Sc	45	41140.8	1.3				ug/L	39543	Standard
	Ti	47	371.0	4.6	**1.3119**	0.075	5.7	ug/L	39	Standard
	V	51	355266.5	2.0	**50.3881**	1.016	2.0	ug/L	1415	Standard
	Cr	52	348734.5	0.9	**48.5348**	0.604	1.2	ug/L	6946	Standard
	Cr	53	45906.1	1.2	**51.4332**	1.330	2.6	ug/L	328	Standard
	Mn	55	484874.0	0.4	**73.8916**	1.228	1.7	ug/L	823	Standard
	Co	59	330526.6	1.2	**46.5427**	1.213	2.6	ug/L	252	Standard
	Ni	60	91208.4	1.4	**47.8307**	0.262	0.5	ug/L	192	Standard
	Cu	65	93139.1	0.3	**48.4331**	0.711	1.5	ug/L	141	Standard
	Zn	66	61775.3	0.7	**54.9696**	1.209	2.2	ug/L	491	Standard
>	**Ge**	72	520412.6	1.6				ug/L	515087	Standard
	As	75	57188.6	0.5	**50.3925**	0.724	1.4	ug/L	20	Standard
	Se	82	5643.2	0.1	**48.4036**	0.791	1.6	ug/L	15	Standard
	Se-1	77	3992.2	4.4	**51.7398**	2.452	4.7	ug/L	48	Standard
>	**Ga**	71	230.0	28.3				mg/L	90	Standard
	Rb	85	19777.9	1.1				ug/L	23	Standard
	Y	89	433897.5	1.4				ug/L	448510	Standard
>	**Rh**	103	81.7	39.4				ug/L	15	Standard
	Mo	98	3301.1	2.4	**1.2202**	0.040	3.3	ug/L	11	Standard
	Ag	107	315385.8	0.9	**47.6558**	0.862	1.8	ug/L	134	Standard
	Cd	111	99498.3	0.6	**47.2265**	0.833	1.8	mg/L	7	Standard
	Cd	114	242671.7	0.4	**47.3834**	0.530	1.1	ug/L	16	Standard
>	**In**	115	602380.0	1.2				ug/L	611117	Standard
	Sn	118	866.7	11.7	**0.0467**	0.016	33.5	ug/L	518	Standard
	Sb	123	233832.7	0.4	**50.2714**	0.769	1.5	ug/L	127	Standard
	Ba	135	168472.5	0.5	**82.7716**	1.205	1.5	ug/L	45	Standard
	Ce	140	3190.3	1.8				ug/L	298	Standard
>	**Tb**	159	975538.4	0.7				ug/L	977942	Standard
	Ho	165	55.0	50.6				ug/L	12	Standard
	Tl	203	405391.8	0.6	**48.6809**	0.381	0.8	ug/L	31	Standard
	Tl	205	349844.8	1.2	**48.3796**	0.623	1.3	ug/L	20	Standard
	Pb	206	268733.3	0.7	**50.2550**	0.909	1.8	ug/L	286	Standard
	Pb	207	232854.8	1.2	**48.2799**	1.141	2.4	ug/L	257	Standard
	Pb	208	923192.2	0.6	**49.4693**	0.366	0.7	ug/L	956	Standard
	U	238	355626.8	0.4	**50.2227**	0.621	1.2	ug/L	17	Standard
>	**Bi**	209	550570.6	1.2				ug/L	557080	Standard

Sample ID: L1606115103S WG573751-07
Report Date/Time: Tuesday, June 28, 2016 16:11:24
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	35867.4	1.9	**45.8327**	0.548	1.2	mg/L	32	Standard
K	39	318.3	2.4	**0.8927**	0.038	4.2	mg/L	30	Standard
Ca	43	141.7	5.4	**15.3452**	1.333	8.7	mg/L	48	Standard
Fe	54	233.9	13.6	**0.0666**	0.029	43.8	mg/L	159	Standard
Fe	57	373.3	5.1	**0.5873**	0.073	12.3	mg/L	195	Standard
> **Sc-1**	45	41140.8	1.3				mg/L	39543	Standard
Cl	35	36665.0	1.2				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	38074.5	2.7				ug/L	1647	Standard
P	31	19512.6	0.4				ug/L	15231	Standard
S	34	2426.9	4.9				ug/L	1845	Standard
Sr	88	241.7	11.4				ug/L	70	Standard
C	12	183.3	31.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	66.3	40.2				mg/L	12	Standard
Ho-1	165	55.0	50.6				mg/L	12	Standard
Er	166	76.7	15.1				mg/L	20	Standard
I	127	311628.4	3.8				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		101.197	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		101.034	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606115103S WG573751-07
Report Date/Time: Tuesday, June 28, 2016 16:11:24
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.570
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.831
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115304SD WG573751-08
Sample Date/Time: Tuesday, June 28, 2016 16:12:19
Number of Replicates: 3
Autosampler Position: 236
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82914.6	1.7				ug/L	81760	Standard
	Be	9	51258.3	2.2	**50.8925**	1.250	2.5	ug/L	17	Standard
	Al	27	15399778.5	0.9	**98.5192**	1.923	2.0	ug/L	1410	Standard
	Sc	45	40084.7	2.9				ug/L	39543	Standard
	Ti	47	640.0	9.7	**2.4715**	0.230	9.3	ug/L	39	Standard
	V	51	358753.7	1.2	**52.7008**	0.680	1.3	ug/L	1415	Standard
	Cr	52	354784.4	1.1	**51.1849**	0.073	0.1	ug/L	6946	Standard
	Cr	53	47661.6	3.5	**55.3029**	1.433	2.6	ug/L	328	Standard
	Mn	55	529314.6	1.5	**83.5469**	1.579	1.9	ug/L	823	Standard
	Co	59	345503.9	1.1	**50.3773**	0.891	1.8	ug/L	252	Standard
	Ni	60	92308.7	1.0	**50.1355**	0.426	0.9	ug/L	192	Standard
	Cu	65	93613.0	1.4	**50.4100**	0.723	1.4	ug/L	141	Standard
	Zn	66	64567.0	0.6	**59.5553**	0.411	0.7	ug/L	491	Standard
>	**Ge**	72	502524.4	1.0				ug/L	515087	Standard
	As	75	56933.1	0.7	**51.9453**	0.162	0.3	ug/L	20	Standard
	Se	82	5947.1	1.7	**52.8275**	0.666	1.3	ug/L	15	Standard
	Se-1	77	3965.5	2.7	**53.2435**	1.801	3.4	ug/L	48	Standard
>	**Ga**	71	291.7	11.2				mg/L	90	Standard
	Rb	85	22576.8	1.0				ug/L	23	Standard
	Y	89	439067.8	2.2				ug/L	448510	Standard
>	**Rh**	103	73.3	7.9				ug/L	15	Standard
	Mo	98	3536.7	2.2	**1.2876**	0.021	1.6	ug/L	11	Standard
	Ag	107	318639.0	0.2	**47.3947**	0.683	1.4	ug/L	134	Standard
	Cd	111	102545.7	0.7	**47.9082**	0.269	0.6	mg/L	7	Standard
	Cd	114	249882.8	0.5	**48.0284**	0.370	0.8	ug/L	16	Standard
>	**In**	115	611940.4	1.2				ug/L	611117	Standard
	Sn	118	1045.0	4.1	**0.0745**	0.005	6.8	ug/L	518	Standard
	Sb	123	235512.7	0.7	**49.8406**	0.693	1.4	ug/L	127	Standard
	Ba	135	174610.3	1.9	**84.4342**	0.545	0.6	ug/L	45	Standard
	Ce	140	5954.5	2.4				ug/L	298	Standard
>	**Tb**	159	973344.6	0.2				ug/L	977942	Standard
	Ho	165	121.7	6.3				ug/L	12	Standard
	Tl	203	413217.7	0.6	**49.7797**	0.456	0.9	ug/L	31	Standard
	Tl	205	363310.0	1.7	**50.4036**	0.986	2.0	ug/L	20	Standard
	Pb	206	270026.6	0.5	**50.6539**	0.212	0.4	ug/L	286	Standard
	Pb	207	235774.0	0.4	**49.0358**	0.293	0.6	ug/L	257	Standard
	Pb	208	947837.4	0.9	**50.9548**	0.634	1.2	ug/L	956	Standard
	U	238	363550.4	1.8	**51.5051**	1.028	2.0	ug/L	17	Standard
>	**Bi**	209	548791.6	0.3				ug/L	557080	Standard

Sample ID: L1606115304SD WG573751-08
Report Date/Time: Tuesday, June 28, 2016 16:14:36
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.4900**	18.160	173.1	mg/L	0	Standard
Mg	24	38714.4	0.7	**50.8180**	1.880	3.7	mg/L	32	Standard
K	39	343.3	15.2	**1.0066**	0.203	20.2	mg/L	30	Standard
Ca	43	156.7	1.8	**18.1495**	1.124	6.2	mg/L	48	Standard
Fe	54	374.0	6.7	**0.1967**	0.015	7.7	mg/L	159	Standard
Fe	57	400.0	0.0	**0.7083**	0.040	5.6	mg/L	195	Standard
> **Sc-1**	45	40084.7	2.9				mg/L	39543	Standard
Cl	35	38416.6	0.8				ug/L	27642	Standard
Kr	83	0.3	173.2				ug/L	2	Standard
Br	81	40474.1	6.0				ug/L	1647	Standard
P	31	20383.7	1.9				ug/L	15231	Standard
S	34	2388.5	6.8				ug/L	1845	Standard
Sr	88	235.0	8.5				ug/L	70	Standard
C	12	253.3	25.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	175.2	19.1				mg/L	12	Standard
Ho-1	165	121.7	6.3				mg/L	12	Standard
Er	166	100.0	26.5				mg/L	20	Standard
I	127	354555.7	8.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.412	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.561	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115304SD WG573751-08
Report Date/Time: Tuesday, June 28, 2016 16:14:36
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.135
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.512
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091602 WG573751-01
Sample Date/Time: Tuesday, June 28, 2016 16:15:30
Number of Replicates: 3
Autosampler Position: 237
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83752.6	2.6				ug/L	81760	Standard
	Be	9	50.0	26.5	**0.0395**	0.014	34.7	ug/L	17	Standard
	Al	27	2701508.7	2.5	**17.0954**	0.130	0.8	ug/L	1410	Standard
	Sc	45	39755.5	3.1				ug/L	39543	Standard
	Ti	47	56.0	11.2	**0.0633**	0.027	42.8	ug/L	39	Standard
	V	51	1924.8	3.9	**0.0741**	0.009	12.6	ug/L	1415	Standard
	Cr	52	10285.8	1.8	**0.4408**	0.012	2.7	ug/L	6946	Standard
	Cr	53	1156.7	1.7	**0.9947**	0.041	4.2	ug/L	328	Standard
	Mn	55	253930.8	1.4	**39.5649**	0.851	2.1	ug/L	823	Standard
	Co	59	1951.8	2.9	**0.2502**	0.004	1.8	ug/L	252	Standard
	Ni	60	1859.1	3.7	**0.8956**	0.027	3.0	ug/L	192	Standard
	Cu	65	1000.0	2.6	**0.4428**	0.017	3.9	ug/L	141	Standard
	Zn	66	2280.8	1.9	**1.2819**	0.026	2.1	ug/L	491	Standard
>	**Ge**	72	508105.2	1.4				ug/L	515087	Standard
	As	75	151.5	17.1	**0.1768**	0.022	12.7	ug/L	20	Standard
	Se	82	39.2	12.0	**0.2256**	0.041	18.4	ug/L	15	Standard
	Se-1	77	70.3	2.2	**0.2277**	0.025	11.1	ug/L	48	Standard
>	**Ga**	71	73.3	3.9				mg/L	90	Standard
	Rb	85	9319.5	3.0				ug/L	23	Standard
	Y	89	446629.7	1.4				ug/L	448510	Standard
>	**Rh**	103	35.0	37.8				ug/L	15	Standard
	Mo	98	255.3	4.6	**0.0742**	0.003	4.3	ug/L	11	Standard
	Ag	107	140.0	17.3	**0.0044**	0.004	86.0	ug/L	134	Standard
	Cd	111	34.5	15.3	**0.0130**	0.002	17.2	mg/L	7	Standard
	Cd	114	78.1	40.8	**0.0099**	0.006	59.3	ug/L	16	Standard
>	**In**	115	618896.7	1.2				ug/L	611117	Standard
	Sn	118	878.4	9.4	**0.0447**	0.013	28.2	ug/L	518	Standard
	Sb	123	309.3	20.5	**0.0520**	0.014	26.9	ug/L	127	Standard
	Ba	135	35857.7	1.5	**17.1253**	0.158	0.9	ug/L	45	Standard
	Ce	140	1368.4	3.3				ug/L	298	Standard
>	**Tb**	159	968292.7	1.7				ug/L	977942	Standard
	Ho	165	100.0	15.0				ug/L	12	Standard
	Tl	203	175.7	13.5	**0.0159**	0.003	15.9	ug/L	31	Standard
	Tl	205	158.3	30.0	**0.0192**	0.006	31.5	ug/L	20	Standard
	Pb	206	469.3	2.7	**0.0227**	0.001	2.5	ug/L	286	Standard
	Pb	207	387.3	8.4	**0.0210**	0.006	30.5	ug/L	257	Standard
	Pb	208	1635.4	0.6	**0.0258**	0.001	5.1	ug/L	956	Standard
	U	238	220.3	6.6	**0.0276**	0.001	5.0	ug/L	17	Standard
>	**Bi**	209	560335.5	2.1				ug/L	557080	Standard

Sample ID: L1606091602 WG573751-01
Report Date/Time: Tuesday, June 28, 2016 16:17:47
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**20.5537**	35.591	173.2	mg/L	0	Standard
Mg	24	3380.4	0.3	**4.4213**	0.131	3.0	mg/L	32	Standard
K	39	168.3	9.1	**0.4237**	0.034	8.1	mg/L	30	Standard
Ca	43	108.3	14.8	**11.0098**	2.304	20.9	mg/L	48	Standard
Fe	54	282.4	10.4	**0.1178**	0.034	28.5	mg/L	159	Standard
Fe	57	291.7	21.2	**0.3561**	0.217	61.0	mg/L	195	Standard
> **Sc-1**	45	39755.5	3.1				mg/L	39543	Standard
Cl	35	40081.0	2.5				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	4644.1	13.8				ug/L	1647	Standard
P	31	18102.5	2.7				ug/L	15231	Standard
S	34	2331.8	2.8				ug/L	1845	Standard
Sr	88	118.3	8.8				ug/L	70	Standard
C	12	123.3	9.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	117.9	56.1				mg/L	12	Standard
Ho-1	165	100.0	15.0				mg/L	12	Standard
Er	166	43.3	48.0				mg/L	20	Standard
I	127	21574.4	25.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.437	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.645	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091602 WG573751-01
Report Date/Time: Tuesday, June 28, 2016 16:17:47
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.273
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.584
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091602DP WG573751-06
Sample Date/Time: Tuesday, June 28, 2016 16:18:41
Number of Replicates: 3
Autosampler Position: 238
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81912.6	2.5				ug/L	81760	Standard
	Be	9	38.3	32.8	**0.0286**	0.012	41.1	ug/L	17	Standard
	Al	27	2661315.3	2.5	**17.2245**	0.558	3.2	ug/L	1410	Standard
	Sc	45	39553.2	2.2				ug/L	39543	Standard
	Ti	47	60.3	10.8	**0.0810**	0.030	36.5	ug/L	39	Standard
	V	51	1771.3	1.7	**0.0516**	0.003	5.3	ug/L	1415	Standard
	Cr	52	10216.4	1.4	**0.4303**	0.018	4.2	ug/L	6946	Standard
	Cr	53	1328.4	4.4	**1.1923**	0.072	6.1	ug/L	328	Standard
	Mn	55	248102.9	0.7	**38.6399**	0.746	1.9	ug/L	823	Standard
	Co	59	1844.8	1.4	**0.2347**	0.001	0.3	ug/L	252	Standard
	Ni	60	1828.1	4.4	**0.8792**	0.056	6.4	ug/L	192	Standard
	Cu	65	995.0	2.4	**0.4399**	0.012	2.7	ug/L	141	Standard
	Zn	66	2165.5	0.6	**1.1746**	0.022	1.8	ug/L	491	Standard
>	**Ge**	72	508291.1	1.3				ug/L	515087	Standard
	As	75	90.4	12.4	**0.1217**	0.009	7.5	ug/L	20	Standard
	Se	82	34.1	11.8	**0.1803**	0.032	17.6	ug/L	15	Standard
	Se-1	77	74.3	14.5	**0.2819**	0.154	54.5	ug/L	48	Standard
>	**Ga**	71	68.3	30.5				mg/L	90	Standard
	Rb	85	9142.8	6.3				ug/L	23	Standard
	Y	89	450226.3	2.8				ug/L	448510	Standard
>	**Rh**	103	21.7	58.1				ug/L	15	Standard
	Mo	98	245.2	4.8	**0.0704**	0.005	7.4	ug/L	11	Standard
	Ag	107	87.7	6.3	**-0.0033**	0.001	21.6	ug/L	134	Standard
	Cd	111	23.2	17.5	**0.0078**	0.002	23.8	mg/L	7	Standard
	Cd	114	58.3	13.8	**0.0061**	0.001	23.6	ug/L	16	Standard
>	**In**	115	619551.5	1.3				ug/L	611117	Standard
	Sn	118	933.4	2.4	**0.0538**	0.006	10.7	ug/L	518	Standard
	Sb	123	201.2	19.7	**0.0293**	0.009	30.0	ug/L	127	Standard
	Ba	135	34661.9	0.9	**16.5369**	0.177	1.1	ug/L	45	Standard
	Ce	140	1260.1	4.6				ug/L	298	Standard
>	**Tb**	159	963340.8	1.2				ug/L	977942	Standard
	Ho	165	88.3	18.2				ug/L	12	Standard
	Tl	203	85.3	6.5	**0.0054**	0.001	13.0	ug/L	31	Standard
	Tl	205	83.3	17.3	**0.0091**	0.002	21.1	ug/L	20	Standard
	Pb	206	415.0	2.1	**0.0134**	0.001	8.9	ug/L	286	Standard
	Pb	207	359.0	2.4	**0.0158**	0.001	8.8	ug/L	257	Standard
	Pb	208	1485.0	2.9	**0.0185**	0.003	14.1	ug/L	956	Standard
	U	238	164.0	3.4	**0.0200**	0.001	3.7	ug/L	17	Standard
>	**Bi**	209	555403.6	0.6				ug/L	557080	Standard

Sample ID: L1606091602DP WG573751-06
Report Date/Time: Tuesday, June 28, 2016 16:20:58
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3443.7	5.9	**4.5314**	0.377	8.3	mg/L	32	Standard
K	39	158.3	7.9	**0.3934**	0.041	10.4	mg/L	30	Standard
Ca	43	93.3	12.4	**8.8061**	1.479	16.8	mg/L	48	Standard
Fe	54	304.0	6.5	**0.1380**	0.017	12.0	mg/L	159	Standard
Fe	57	313.3	6.4	**0.4340**	0.087	20.1	mg/L	195	Standard
> **Sc-1**	45	39553.2	2.2				mg/L	39543	Standard
Cl	35	40893.8	1.8				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	3857.2	4.5				ug/L	1647	Standard
P	31	18868.4	1.0				ug/L	15231	Standard
S	34	2383.5	7.0				ug/L	1845	Standard
Sr	88	95.0	13.9				ug/L	70	Standard
C	12	200.0	17.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	96.2	30.5				mg/L	12	Standard
Ho-1	165	88.3	18.2				mg/L	12	Standard
Er	166	80.0	37.5				mg/L	20	Standard
I	127	15653.1	0.7				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		100.187	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		98.681	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606091602DP WG573751-06
Report Date/Time: Tuesday, June 28, 2016 16:20:58
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	101.380	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	99.699	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091601
Sample Date/Time: Tuesday, June 28, 2016 16:21:53
Number of Replicates: 3
Autosampler Position: 239
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81664.5	1.8				ug/L	81760	Standard
	Be	9	26.7	28.6	**0.0172**	0.008	47.1	ug/L	17	Standard
	Al	27	4275643.4	2.5	**27.7561**	0.463	1.7	ug/L	1410	Standard
	Sc	45	38087.8	3.0				ug/L	39543	Standard
	Ti	47	54.7	26.0	**0.0548**	0.059	108.3	ug/L	39	Standard
	V	51	1697.6	8.1	**0.0374**	0.021	56.8	ug/L	1415	Standard
	Cr	52	10296.2	0.8	**0.4201**	0.020	4.7	ug/L	6946	Standard
	Cr	53	1460.1	7.5	**1.3201**	0.135	10.2	ug/L	328	Standard
	Mn	55	64493.4	3.0	**9.7887**	0.389	4.0	ug/L	823	Standard
	Co	59	427.0	5.1	**0.0294**	0.004	12.3	ug/L	252	Standard
	Ni	60	847.0	2.5	**0.3443**	0.016	4.7	ug/L	192	Standard
	Cu	65	826.7	6.2	**0.3435**	0.022	6.5	ug/L	141	Standard
	Zn	66	9487.6	0.7	**7.8171**	0.052	0.7	ug/L	491	Standard
>	**Ge**	72	515789.8	1.2				ug/L	515087	Standard
	As	75	95.2	47.1	**0.1249**	0.040	32.1	ug/L	20	Standard
	Se	82	32.5	28.2	**0.1620**	0.079	48.7	ug/L	15	Standard
	Se-1	77	78.3	8.3	**0.3202**	0.098	30.6	ug/L	48	Standard
>	**Ga**	71	58.3	50.2				mg/L	90	Standard
	Rb	85	8367.3	1.9				ug/L	23	Standard
	Y	89	424791.4	2.4				ug/L	448510	Standard
>	**Rh**	103	81.7	19.7				ug/L	15	Standard
	Mo	98	367.2	2.9	**0.1192**	0.004	3.5	ug/L	11	Standard
	Ag	107	119.3	48.0	**0.0019**	0.009	460.3	ug/L	134	Standard
	Cd	111	50.7	68.6	**0.0212**	0.017	78.0	mg/L	7	Standard
	Cd	114	102.4	71.8	**0.0151**	0.014	95.2	ug/L	16	Standard
>	**In**	115	600349.6	1.0				ug/L	611117	Standard
	Sn	118	1276.7	2.5	**0.1178**	0.004	3.0	ug/L	518	Standard
	Sb	123	200.9	15.6	**0.0306**	0.007	23.5	ug/L	127	Standard
	Ba	135	168683.5	0.8	**83.1505**	0.843	1.0	ug/L	45	Standard
	Ce	140	160.0	17.4				ug/L	298	Standard
>	**Tb**	159	943339.1	0.8				ug/L	977942	Standard
	Ho	165	5.0	100.0				ug/L	12	Standard
	Tl	203	103.7	83.0	**0.0076**	0.010	135.3	ug/L	31	Standard
	Tl	205	65.0	38.5	**0.0066**	0.003	51.6	ug/L	20	Standard
	Pb	206	449.7	4.5	**0.0200**	0.004	18.4	ug/L	286	Standard
	Pb	207	373.7	5.2	**0.0190**	0.004	20.9	ug/L	257	Standard
	Pb	208	1469.0	1.7	**0.0178**	0.001	8.0	ug/L	956	Standard
	U	238	195.0	9.9	**0.0244**	0.003	11.1	ug/L	17	Standard
>	**Bi**	209	554105.5	0.2				ug/L	557080	Standard

Sample ID: L1606091601
Report Date/Time: Tuesday, June 28, 2016 16:24:10
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**11.2769**	19.523	173.1	mg/L	0	Standard
Mg	24	4872.5	3.1	**6.6831**	0.365	5.5	mg/L	32	Standard
K	39	180.0	11.1	**0.4891**	0.054	11.0	mg/L	30	Standard
Ca	43	71.7	49.5	**5.8566**	5.217	89.1	mg/L	48	Standard
Fe	54	187.0	32.7	**0.0391**	0.060	153.5	mg/L	159	Standard
Fe	57	275.0	13.7	**0.3422**	0.158	46.1	mg/L	195	Standard
> **Sc-1**	45	38087.8	3.0				mg/L	39543	Standard
Cl	35	39985.4	2.3				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	5007.5	9.4				ug/L	1647	Standard
P	31	18170.9	2.8				ug/L	15231	Standard
S	34	2413.5	5.6				ug/L	1845	Standard
Sr	88	318.3	14.3				ug/L	70	Standard
C	12	223.3	5.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	59.7	16.8				mg/L	12	Standard
Ho-1	165	5.0	100.0				mg/L	12	Standard
Er	166	6.7	86.6				mg/L	20	Standard
I	127	27700.4	6.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		99.883	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.136	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091601
Report Date/Time: Tuesday, June 28, 2016 16:24:10
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.238
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.466
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091601PS WG574522-01
Sample Date/Time: Tuesday, June 28, 2016 16:25:04
Number of Replicates: 3
Autosampler Position: 240
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84128.0	1.7				ug/L	81760	Standard
	Be	9	52619.6	0.9	**51.4982**	1.298	2.5	ug/L	17	Standard
	Al	27	4326633.2	0.4	**27.2707**	0.550	2.0	ug/L	1410	Standard
	Sc	45	39650.1	0.7				ug/L	39543	Standard
	Ti	47	62.3	14.0	**0.0872**	0.038	43.9	ug/L	39	Standard
	V	51	353435.0	1.4	**50.9243**	0.391	0.8	ug/L	1415	Standard
	Cr	52	362779.9	0.9	**51.3507**	0.534	1.0	ug/L	6946	Standard
	Cr	53	45635.3	3.2	**51.9244**	0.938	1.8	ug/L	328	Standard
	Mn	55	384138.6	0.3	**59.4407**	0.957	1.6	ug/L	823	Standard
	Co	59	352427.3	0.4	**50.4121**	0.778	1.5	ug/L	252	Standard
	Ni	60	95802.0	1.3	**51.0471**	0.630	1.2	ug/L	192	Standard
	Cu	65	98413.6	0.9	**51.9931**	0.640	1.2	ug/L	141	Standard
	Zn	66	63162.1	1.4	**57.1168**	0.348	0.6	ug/L	491	Standard
>	**Ge**	72	512293.3	1.9				ug/L	515087	Standard
	As	75	56566.0	1.5	**50.6330**	0.812	1.6	ug/L	20	Standard
	Se	82	5805.6	1.0	**50.5901**	0.677	1.3	ug/L	15	Standard
	Se-1	77	3921.2	1.4	**51.6223**	0.909	1.8	ug/L	48	Standard
>	**Ga**	71	60.0	50.7				mg/L	90	Standard
	Rb	85	8215.6	0.6				ug/L	23	Standard
	Y	89	438501.2	1.8				ug/L	448510	Standard
>	**Rh**	103	71.7	28.2				ug/L	15	Standard
	Mo	98	368.8	5.9	**0.1169**	0.008	6.7	ug/L	11	Standard
	Ag	107	334626.2	0.5	**49.6506**	0.360	0.7	ug/L	134	Standard
	Cd	111	105047.6	0.7	**48.9601**	0.529	1.1	mg/L	7	Standard
	Cd	114	254060.7	2.8	**48.7162**	1.522	3.1	ug/L	16	Standard
>	**In**	115	613389.4	0.3				ug/L	611117	Standard
	Sn	118	1936.8	37.5	**0.2241**	0.121	54.0	ug/L	518	Standard
	Sb	123	237358.9	1.9	**50.1080**	0.980	2.0	ug/L	127	Standard
	Ba	135	267735.0	1.3	**129.1838**	2.081	1.6	ug/L	45	Standard
	Ce	140	231.7	21.3				ug/L	298	Standard
>	**Tb**	159	964479.2	1.1				ug/L	977942	Standard
	Ho	165	18.3	56.8				ug/L	12	Standard
	Tl	203	432424.0	0.5	**50.0145**	0.262	0.5	ug/L	31	Standard
	Tl	205	382714.9	0.3	**50.9759**	0.183	0.4	ug/L	20	Standard
	Pb	206	282966.8	1.1	**50.9646**	0.635	1.2	ug/L	286	Standard
	Pb	207	246387.7	1.3	**49.1992**	0.690	1.4	ug/L	257	Standard
	Pb	208	985942.4	0.8	**50.8876**	0.399	0.8	ug/L	956	Standard
	U	238	365334.3	1.0	**49.6910**	0.475	1.0	ug/L	17	Standard
>	**Bi**	209	571593.5	0.1				ug/L	557080	Standard

Sample ID: L1606091601PS WG574522-01
Report Date/Time: Tuesday, June 28, 2016 16:27:21
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4890.8	3.1	**6.4355**	0.187	2.9	mg/L	32	Standard
K	39	161.7	23.2	**0.4031**	0.125	30.9	mg/L	30	Standard
Ca	43	113.3	24.3	**11.8087**	4.086	34.6	mg/L	48	Standard
Fe	54	177.2	7.7	**0.0228**	0.011	49.5	mg/L	159	Standard
Fe	57	255.0	15.6	**0.2338**	0.132	56.3	mg/L	195	Standard
> **Sc-1**	45	39650.1	0.7				mg/L	39543	Standard
Cl	35	40732.7	2.1				ug/L	27642	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	4914.1	11.2				ug/L	1647	Standard
P	31	18529.7	1.8				ug/L	15231	Standard
S	34	2476.9	6.8				ug/L	1845	Standard
Sr	88	310.0	4.8				ug/L	70	Standard
C	12	160.0	18.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	12.7	117.1				mg/L	12	Standard
Ho-1	165	18.3	56.8				mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	27421.5	4.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.896	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.458	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091601PS WG574522-01
Report Date/Time: Tuesday, June 28, 2016 16:27:21
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.372
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.605
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Ba 135 Upper, S, EEE	Ba	135	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091601SDL WG574522-02
Sample Date/Time: Tuesday, June 28, 2016 16:28:16
Number of Replicates: 3
Autosampler Position: 241
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82723.6	2.3				ug/L	81760	Standard
	Be	9	23.3	12.4	**0.0135**	0.003	18.7	ug/L	17	Standard
	Al	27	900313.9	2.1	**5.7596**	0.136	2.4	ug/L	1410	Standard
	Sc	45	39763.8	0.8				ug/L	39543	Standard
	Ti	47	29.7	9.7	**-0.0430**	0.012	28.1	ug/L	39	Standard
	V	51	1735.9	7.6	**0.0491**	0.020	40.6	ug/L	1415	Standard
	Cr	52	8352.3	2.2	**0.1715**	0.035	20.7	ug/L	6946	Standard
	Cr	53	645.0	2.8	**0.4109**	0.024	5.9	ug/L	328	Standard
	Mn	55	14377.5	0.8	**2.1221**	0.018	0.8	ug/L	823	Standard
	Co	59	301.0	11.3	**0.0125**	0.005	41.5	ug/L	252	Standard
	Ni	60	410.7	1.2	**0.1183**	0.004	3.1	ug/L	192	Standard
	Cu	65	611.3	5.6	**0.2385**	0.017	7.2	ug/L	141	Standard
	Zn	66	7975.8	0.8	**6.6226**	0.085	1.3	ug/L	491	Standard
>	**Ge**	72	503109.1	0.7				ug/L	515087	Standard
	As	75	19.7	109.3	**0.0582**	0.020	33.7	ug/L	20	Standard
	Se	82	17.8	58.3	**0.0388**	0.092	237.9	ug/L	15	Standard
	Se-1	77	57.7	14.7	**0.0653**	0.120	183.5	ug/L	48	Standard
>	**Ga**	71	58.3	44.0				mg/L	90	Standard
	Rb	85	1753.4	3.2				ug/L	23	Standard
	Y	89	444634.9	0.6				ug/L	448510	Standard
>	**Rh**	103	21.7	48.0				ug/L	15	Standard
	Mo	98	88.5	3.1	**0.0131**	0.001	5.4	ug/L	11	Standard
	Ag	107	142.0	4.2	**0.0046**	0.001	16.3	ug/L	134	Standard
	Cd	111	23.2	23.8	**0.0077**	0.002	31.6	mg/L	7	Standard
	Cd	114	28.0	34.6	**0.0004**	0.002	477.3	ug/L	16	Standard
>	**In**	115	620921.1	1.0				ug/L	611117	Standard
	Sn	118	821.7	5.2	**0.0348**	0.006	16.8	ug/L	518	Standard
	Sb	123	2051.2	21.4	**0.4156**	0.096	23.0	ug/L	127	Standard
	Ba	135	34796.2	0.2	**16.5640**	0.119	0.7	ug/L	45	Standard
	Ce	140	63.3	12.1				ug/L	298	Standard
>	**Tb**	159	973024.2	1.8				ug/L	977942	Standard
	Ho	165	8.3	124.9				ug/L	12	Standard
	Tl	203	97.0	5.7	**0.0066**	0.000	6.5	ug/L	31	Standard
	Tl	205	68.3	44.1	**0.0070**	0.004	57.1	ug/L	20	Standard
	Pb	206	360.7	11.0	**0.0024**	0.006	236.3	ug/L	286	Standard
	Pb	207	322.3	9.7	**0.0075**	0.006	86.2	ug/L	257	Standard
	Pb	208	1339.4	1.1	**0.0099**	0.001	10.7	ug/L	956	Standard
	U	238	86.3	6.6	**0.0090**	0.001	5.9	ug/L	17	Standard
>	**Bi**	209	562189.7	2.5				ug/L	557080	Standard

Sample ID: L1606091601SDL WG574522-02
Report Date/Time: Tuesday, June 28, 2016 16:30:33
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	991.7	7.3	**1.2554**	0.092	7.3	mg/L	32	Standard
K	39	51.7	11.2	**0.0323**	0.021	64.1	mg/L	30	Standard
Ca	43	40.0	45.1	**0.6884**	2.694	391.4	mg/L	48	Standard
Fe	54	140.9	19.7	**-0.0104**	0.024	229.0	mg/L	159	Standard
Fe	57	241.7	4.8	**0.1865**	0.032	17.2	mg/L	195	Standard
> **Sc-1**	45	39763.8	0.8				mg/L	39543	Standard
Cl	35	40635.1	0.7				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	2310.2	7.5				ug/L	1647	Standard
P	31	19332.3	4.6				ug/L	15231	Standard
S	34	2451.9	6.9				ug/L	1845	Standard
Sr	88	125.0	21.2				ug/L	70	Standard
C	12	163.3	25.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	19.5	88.8				mg/L	12	Standard
Ho-1	165	8.3	124.9				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	7598.6	6.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.179	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.675	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091601SDL WG574522-02
Report Date/Time: Tuesday, June 28, 2016 16:30:33
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.604
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.917
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 16:31:29
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	79881.8	2.3				ug/L	81760	Standard
	Be	9	50840.2	1.3	**52.4015**	1.145	2.2	ug/L	17	Standard
	Al	27	7585830.9	1.9	**50.3677**	1.194	2.4	ug/L	1410	Standard
	Sc	45	39128.8	0.6				ug/L	39543	Standard
	Ti	47	24998.6	0.9	**102.1788**	1.762	1.7	ug/L	39	Standard
	V	51	348429.9	1.4	**50.8229**	0.271	0.5	ug/L	1415	Standard
	Cr	52	355416.3	0.4	**50.9232**	0.579	1.1	ug/L	6946	Standard
	Cr	53	44443.4	1.1	**51.2031**	0.906	1.8	ug/L	328	Standard
	Mn	55	315927.2	0.8	**49.4624**	0.575	1.2	ug/L	823	Standard
	Co	59	353189.3	0.9	**51.1457**	0.854	1.7	ug/L	252	Standard
	Ni	60	94600.8	0.9	**51.0293**	0.136	0.3	ug/L	192	Standard
	Cu	65	95833.5	0.4	**51.2541**	0.406	0.8	ug/L	141	Standard
	Zn	66	56109.7	0.5	**51.2906**	0.852	1.7	ug/L	491	Standard
>	Ge	72	505999.1	1.2				ug/L	515087	Standard
	As	75	55945.5	0.4	**50.6968**	0.402	0.8	ug/L	20	Standard
	Se	82	5756.3	0.8	**50.7852**	1.001	2.0	ug/L	15	Standard
	Se-1	77	3889.8	1.1	**51.8484**	1.021	2.0	ug/L	48	Standard
>	Ga	71	51.7	20.1				mg/L	90	Standard
	Rb	85	811.7	9.0				ug/L	23	Standard
	Y	89	426423.1	2.5				ug/L	448510	Standard
>	Rh	103	31.7	24.1				ug/L	15	Standard
	Mo	98	265386.1	0.2	**99.3199**	0.619	0.6	ug/L	11	Standard
	Ag	107	333927.0	1.2	**50.3137**	0.313	0.6	ug/L	134	Standard
	Cd	111	105484.6	1.1	**49.9277**	0.764	1.5	mg/L	7	Standard
	Cd	114	258596.4	2.6	**50.3494**	1.108	2.2	ug/L	16	Standard
>	In	115	604027.1	0.8				ug/L	611117	Standard
	Sn	118	295796.4	1.2	**50.5150**	0.923	1.8	ug/L	518	Standard
	Sb	123	237014.9	1.1	**50.8107**	0.333	0.7	ug/L	127	Standard
	Ba	135	100914.0	1.4	**49.4318**	0.852	1.7	ug/L	45	Standard
	Ce	140	80.0	0.0				ug/L	298	Standard
>	Tb	159	968181.4	1.1				ug/L	977942	Standard
	Ho	165	8.3	69.3				ug/L	12	Standard
	Tl	203	430250.2	1.2	**50.7804**	1.115	2.2	ug/L	31	Standard
	Tl	205	381434.8	2.3	**51.8414**	1.366	2.6	ug/L	20	Standard
	Pb	206	273103.8	0.8	**50.1867**	0.395	0.8	ug/L	286	Standard
	Pb	207	246704.6	0.5	**50.2667**	0.621	1.2	ug/L	257	Standard
	Pb	208	974193.4	0.6	**51.3084**	0.901	1.8	ug/L	956	Standard
	U	238	366221.1	0.4	**50.8280**	0.746	1.5	ug/L	17	Standard
>	Bi	209	560252.8	1.6				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 16:33:46
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3943.8	5.4	**5.2490**	0.301	5.7	mg/L	32	Standard
K	39	1356.7	5.3	**4.4904**	0.223	5.0	mg/L	30	Standard
Ca	43	83.3	18.3	**7.4389**	2.295	30.8	mg/L	48	Standard
Fe	54	5957.3	2.0	**5.3112**	0.093	1.8	mg/L	159	Standard
Fe	57	1600.1	10.0	**4.8365**	0.580	12.0	mg/L	195	Standard
> **Sc-1**	45	39128.8	0.6				mg/L	39543	Standard
Cl	35	37049.3	3.3				ug/L	27642	Standard
Kr	83	2.0	132.3				ug/L	2	Standard
Br	81	1520.1	14.1				ug/L	1647	Standard
P	31	19898.1	2.1				ug/L	15231	Standard
S	34	2718.6	2.9				ug/L	1845	Standard
Sr	88	73.3	20.8				ug/L	70	Standard
C	12	120.0	43.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	9.0	113.3				mg/L	12	Standard
Ho-1	165	8.3	69.3				mg/L	12	Standard
Er	166	20.0	50.0				mg/L	20	Standard
I	127	2098.5	16.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	104.803		
Al	27	100.735		
Sc	45			
Ti	47	102.179		
V	51	101.646		
Cr	52	101.846		
Cr	53			
Mn	55	98.925		
Co	59	102.291		
Ni	60	102.059		
Cu	65	102.508		
Zn	66	102.581		
> Ge	72		98.236	
As	75	101.394		
Se	82	101.570		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 16:33:46
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	99.320		
Ag	107	100.627		
Cd	111	99.855		
Cd	114			
> In	115			98.840
Sn	118	101.030		
Sb	123	101.621		
Ba	135	98.864		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.561		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.617		
U	238	101.656		
> Bi	209			100.569
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 16:33:46
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 16:34:41
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77653.6	3.2				ug/L	81760	Standard
	Be	9	28.3	27.0	**0.0205**	0.009	43.8	ug/L	17	Standard
	Al	27	2033.5	6.8	**-0.0005**	0.001	113.9	ug/L	1410	Standard
	Sc	45	38149.6	0.7				ug/L	39543	Standard
	Ti	47	33.0	16.0	**-0.0290**	0.023	78.1	ug/L	39	Standard
	V	51	1124.7	5.5	**-0.0407**	0.010	23.8	ug/L	1415	Standard
	Cr	52	6085.9	0.6	**-0.1604**	0.016	10.1	ug/L	6946	Standard
	Cr	53	345.0	11.5	**0.0612**	0.041	67.1	ug/L	328	Standard
	Mn	55	652.0	2.2	**-0.0453**	0.002	5.1	ug/L	823	Standard
	Co	59	236.0	7.4	**0.0031**	0.003	92.5	ug/L	252	Standard
	Ni	60	163.3	9.5	**-0.0160**	0.008	51.9	ug/L	192	Standard
	Cu	65	132.3	11.7	**-0.0193**	0.009	47.2	ug/L	141	Standard
	Zn	66	446.3	5.3	**-0.4102**	0.017	4.3	ug/L	491	Standard
>	**Ge**	72	502289.8	1.3				ug/L	515087	Standard
	As	75	-42.3	118.5	**0.0014**	0.046	3304.8	ug/L	20	Standard
	Se	82	15.3	48.7	**0.0168**	0.064	382.8	ug/L	15	Standard
	Se-1	77	51.0	21.6	**-0.0257**	0.142	551.0	ug/L	48	Standard
>	**Ga**	71	38.3	27.2				mg/L	90	Standard
	Rb	85	10.0	86.6				ug/L	23	Standard
	Y	89	438316.4	0.6				ug/L	448510	Standard
>	**Rh**	103	15.0	0.0				ug/L	15	Standard
	Mo	98	228.3	18.4	**0.0667**	0.015	22.0	ug/L	11	Standard
	Ag	107	123.7	9.2	**0.0025**	0.002	72.5	ug/L	134	Standard
	Cd	111	11.3	56.7	**0.0024**	0.003	123.9	mg/L	7	Standard
	Cd	114	32.3	39.0	**0.0014**	0.003	178.7	ug/L	16	Standard
>	**In**	115	601315.6	1.3				ug/L	611117	Standard
	Sn	118	783.4	33.2	**0.0331**	0.047	140.5	ug/L	518	Standard
	Sb	123	646.2	31.1	**0.1267**	0.045	35.7	ug/L	127	Standard
	Ba	135	40.3	7.6	**-0.0052**	0.002	33.4	ug/L	45	Standard
	Ce	140	38.3	19.9				ug/L	298	Standard
>	**Tb**	159	937082.2	1.4				ug/L	977942	Standard
	Ho	165	5.0	100.0				ug/L	12	Standard
	Tl	203	49.7	12.9	**0.0012**	0.001	66.7	ug/L	31	Standard
	Tl	205	40.0	57.3	**0.0033**	0.003	97.6	ug/L	20	Standard
	Pb	206	275.3	5.0	**-0.0116**	0.003	25.8	ug/L	286	Standard
	Pb	207	258.0	7.0	**-0.0041**	0.003	84.6	ug/L	257	Standard
	Pb	208	1051.3	3.4	**-0.0036**	0.002	41.9	ug/L	956	Standard
	U	238	50.3	16.2	**0.0042**	0.001	28.1	ug/L	17	Standard
>	**Bi**	209	546028.7	0.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 16:36:58
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**22.1550**	19.184	86.6	mg/L	0	Standard
Mg	24	31.7	50.8	**-0.0134**	0.022	167.6	mg/L	32	Standard
K	39	30.0	28.9	**-0.0362**	0.031	85.5	mg/L	30	Standard
Ca	43	30.0	66.7	**-0.6084**	3.177	522.1	mg/L	48	Standard
Fe	54	168.0	24.0	**0.0206**	0.039	187.8	mg/L	159	Standard
Fe	57	143.3	24.5	**-0.1229**	0.124	101.2	mg/L	195	Standard
Sc-1	45	38149.6	0.7				mg/L	39543	Standard
Cl	35	36673.0	2.0				ug/L	27642	Standard
Kr	83	1.0	173.2				ug/L	2	Standard
Br	81	1570.1	7.0				ug/L	1647	Standard
P	31	19050.3	3.5				ug/L	15231	Standard
S	34	2736.9	0.8				ug/L	1845	Standard
Sr	88	48.3	15.8				ug/L	70	Standard
C	12	133.3	4.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	5.7	217.0				mg/L	12	Standard
Ho-1	165	5.0	100.0				mg/L	12	Standard
Er	166	20.0	100.0				mg/L	20	Standard
I	127	1683.4	18.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		97.516	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 16:36:58
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.396
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.016
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606091601SDL WG574522-02
Sample Date/Time: Tuesday, June 28, 2016 16:37:54
Number of Replicates: 3
Autosampler Position: 242
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80481.5	1.5				ug/L	81760	Standard
	Be	9	31.7	9.1	**0.0226**	0.003	11.9	ug/L	17	Standard
	Al	27	181010.0	1.7	**1.1789**	0.039	3.3	ug/L	1410	Standard
	Sc	45	40285.2	2.7				ug/L	39543	Standard
	Ti	47	30.3	10.1	**-0.0420**	0.014	34.3	ug/L	39	Standard
	V	51	1262.3	9.3	**-0.0236**	0.014	61.0	ug/L	1415	Standard
	Cr	52	6637.8	2.8	**-0.0958**	0.014	14.1	ug/L	6946	Standard
	Cr	53	403.3	17.4	**0.1211**	0.074	61.5	ug/L	328	Standard
	Mn	55	3292.4	3.2	**0.3636**	0.014	3.8	ug/L	823	Standard
	Co	59	275.7	13.5	**0.0082**	0.005	64.5	ug/L	252	Standard
	Ni	60	218.7	22.1	**0.0120**	0.025	207.8	ug/L	192	Standard
	Cu	65	189.7	2.7	**0.0098**	0.003	33.4	ug/L	141	Standard
	Zn	66	2022.1	5.3	**1.0328**	0.100	9.7	ug/L	491	Standard
>	**Ge**	72	510876.9	1.9				ug/L	515087	Standard
	As	75	-16.1	411.9	**0.0255**	0.059	233.0	ug/L	20	Standard
	Se	82	17.9	79.1	**0.0373**	0.123	328.7	ug/L	15	Standard
	Se-1	77	53.0	5.0	**-0.0093**	0.040	429.4	ug/L	48	Standard
>	**Ga**	71	26.7	10.8				mg/L	90	Standard
	Rb	85	385.0	11.5				ug/L	23	Standard
	Y	89	434970.9	1.3				ug/L	448510	Standard
>	**Rh**	103	13.3	57.3				ug/L	15	Standard
	Mo	98	105.1	23.0	**0.0193**	0.009	45.6	ug/L	11	Standard
	Ag	107	160.3	33.6	**0.0074**	0.008	107.2	ug/L	134	Standard
	Cd	111	27.8	77.7	**0.0099**	0.010	100.3	mg/L	7	Standard
	Cd	114	89.4	76.6	**0.0121**	0.013	107.6	ug/L	16	Standard
>	**In**	115	618555.8	0.1				ug/L	611117	Standard
	Sn	118	705.0	14.2	**0.0159**	0.017	105.2	ug/L	518	Standard
	Sb	123	323.1	28.2	**0.0548**	0.019	34.7	ug/L	127	Standard
	Ba	135	6967.6	1.3	**3.3092**	0.040	1.2	ug/L	45	Standard
	Ce	140	23.3	65.5				ug/L	298	Standard
>	**Tb**	159	976647.8	1.4				ug/L	977942	Standard
	Ho	165	15.0	57.7				ug/L	12	Standard
	Tl	203	112.7	67.1	**0.0081**	0.009	105.4	ug/L	31	Standard
	Tl	205	90.0	29.4	**0.0096**	0.003	35.2	ug/L	20	Standard
	Pb	206	345.7	14.3	**-0.0019**	0.008	437.8	ug/L	286	Standard
	Pb	207	297.3	11.3	**0.0008**	0.006	736.1	ug/L	257	Standard
	Pb	208	1206.4	8.6	**0.0013**	0.005	370.6	ug/L	956	Standard
	U	238	79.3	55.4	**0.0077**	0.006	75.6	ug/L	17	Standard
>	**Bi**	209	576852.7	1.1				ug/L	557080	Standard

Sample ID: L1606091601SDL WG574522-02
Report Date/Time: Tuesday, June 28, 2016 16:40:10
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	251.7	21.2	**0.2729**	0.077	28.4	mg/L	32	Standard
K	39	35.0	75.6	**-0.0238**	0.089	375.8	mg/L	30	Standard
Ca	43	41.7	66.1	**0.8623**	4.157	482.1	mg/L	48	Standard
Fe	54	139.7	6.2	**-0.0128**	0.010	79.6	mg/L	159	Standard
Fe	57	215.0	8.1	**0.0878**	0.054	61.5	mg/L	195	Standard
> **Sc-1**	45	40285.2	2.7				mg/L	39543	Standard
Cl	35	36209.9	2.0				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1783.4	3.4				ug/L	1647	Standard
P	31	18659.8	1.9				ug/L	15231	Standard
S	34	2565.2	1.9				ug/L	1845	Standard
Sr	88	76.7	26.4				ug/L	70	Standard
C	12	126.7	24.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	19.2	52.1				mg/L	12	Standard
Ho-1	165	15.0	57.7				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	2670.2	2.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		98.436	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.183	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091601SDL WG574522-02
Report Date/Time: Tuesday, June 28, 2016 16:40:10
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.217
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.549
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606114201
Sample Date/Time: Tuesday, June 28, 2016 16:41:05
Number of Replicates: 3
Autosampler Position: 243
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80520.1	1.9				ug/L	81760	Standard
	Be	9	30.0	16.7	**0.0209**	0.005	21.7	ug/L	17	Standard
	Al	27	1224587.7	1.8	**8.0560**	0.286	3.6	ug/L	1410	Standard
	Sc	45	42140.2	1.9				ug/L	39543	Standard
	Ti	47	397.3	21.3	**1.4906**	0.359	24.1	ug/L	39	Standard
	V	51	2122.9	0.3	**0.1097**	0.002	1.5	ug/L	1415	Standard
	Cr	52	14472.9	1.3	**1.0963**	0.024	2.2	ug/L	6946	Standard
	Cr	53	2878.6	1.5	**3.0550**	0.050	1.6	ug/L	328	Standard
	Mn	55	698464.2	0.4	**111.4396**	0.282	0.3	ug/L	823	Standard
	Co	59	22132.5	0.4	**3.2312**	0.012	0.4	ug/L	252	Standard
	Ni	60	7435.8	1.3	**3.9843**	0.043	1.1	ug/L	192	Standard
	Cu	65	724.4	2.6	**0.3040**	0.008	2.7	ug/L	141	Standard
	Zn	66	6121.2	1.0	**4.9565**	0.034	0.7	ug/L	491	Standard
>	**Ge**	72	497323.9	0.5				ug/L	515087	Standard
	As	75	171.3	19.9	**0.1980**	0.031	15.4	ug/L	20	Standard
	Se	82	64.0	10.3	**0.4560**	0.056	12.4	ug/L	15	Standard
	Se-1	77	106.7	7.0	**0.7479**	0.110	14.7	ug/L	48	Standard
>	**Ga**	71	88.3	6.5				mg/L	90	Standard
	Rb	85	4614.0	1.5				ug/L	23	Standard
	Y	89	442742.2	2.0				ug/L	448510	Standard
>	**Rh**	103	21.7	35.3				ug/L	15	Standard
	Mo	98	335.3	5.4	**0.1033**	0.008	7.8	ug/L	11	Standard
	Ag	107	92.0	14.2	**-0.0027**	0.002	71.5	ug/L	134	Standard
	Cd	111	86.4	7.6	**0.0370**	0.004	9.6	mg/L	7	Standard
	Cd	114	204.2	9.6	**0.0339**	0.004	12.0	ug/L	16	Standard
>	**In**	115	619678.1	1.4				ug/L	611117	Standard
	Sn	118	898.4	6.0	**0.0480**	0.011	22.7	ug/L	518	Standard
	Sb	123	324.0	26.0	**0.0551**	0.019	33.6	ug/L	127	Standard
	Ba	135	76873.5	0.9	**36.7017**	0.677	1.8	ug/L	45	Standard
	Ce	140	3832.2	0.5				ug/L	298	Standard
>	**Tb**	159	963958.5	1.3				ug/L	977942	Standard
	Ho	165	96.7	23.3				ug/L	12	Standard
	Tl	203	122.3	9.5	**0.0098**	0.001	14.0	ug/L	31	Standard
	Tl	205	126.7	13.9	**0.0151**	0.002	14.9	ug/L	20	Standard
	Pb	206	564.7	2.8	**0.0417**	0.003	6.6	ug/L	286	Standard
	Pb	207	477.0	3.5	**0.0406**	0.003	8.6	ug/L	257	Standard
	Pb	208	2018.4	2.1	**0.0474**	0.002	3.2	ug/L	956	Standard
	U	238	92.7	4.5	**0.0101**	0.001	6.9	ug/L	17	Standard
>	**Bi**	209	552929.6	1.0				ug/L	557080	Standard

Sample ID: L1606114201
Report Date/Time: Tuesday, June 28, 2016 16:43:22
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.1351**	17.546	173.1	mg/L	0	Standard
Mg	24	7430.2	1.6	**9.2261**	0.222	2.4	mg/L	32	Standard
K	39	131.7	22.3	**0.2750**	0.084	30.6	mg/L	30	Standard
Ca	43	50.0	10.0	**1.7747**	0.605	34.1	mg/L	48	Standard
Fe	54	2749.1	1.6	**2.1977**	0.008	0.4	mg/L	159	Standard
Fe	57	918.4	0.8	**2.2851**	0.038	1.7	mg/L	195	Standard
> **Sc-1**	45	42140.2	1.9				mg/L	39543	Standard
Cl	35	36101.6	0.8				ug/L	27642	Standard
Kr	83	3.3	75.5				ug/L	2	Standard
Br	81	12138.2	7.4				ug/L	1647	Standard
P	31	25736.8	1.2				ug/L	15231	Standard
S	34	2443.5	6.8				ug/L	1845	Standard
Sr	88	88.3	31.2				ug/L	70	Standard
C	12	360.0	10.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	87.0	21.3				mg/L	12	Standard
Ho-1	165	96.7	23.3				mg/L	12	Standard
Er	166	63.3	48.2				mg/L	20	Standard
I	127	223077.6	8.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		98.483	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.551	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606114201
Report Date/Time: Tuesday, June 28, 2016 16:43:22
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	101.401	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	99.255	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606114203
Sample Date/Time: Tuesday, June 28, 2016 16:44:17
Number of Replicates: 3
Autosampler Position: 244
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81992.9	1.1				ug/L	81760	Standard
	Be	9	28.3	27.0	**0.0187**	0.008	41.6	ug/L	17	Standard
	Al	27	794059.8	1.7	**5.1234**	0.139	2.7	ug/L	1410	Standard
	Sc	45	39785.5	1.2				ug/L	39543	Standard
	Ti	47	377.0	7.0	**1.3367**	0.105	7.9	ug/L	39	Standard
	V	51	6234.1	1.0	**0.6819**	0.005	0.8	ug/L	1415	Standard
	Cr	52	12611.2	0.3	**0.7377**	0.006	0.8	ug/L	6946	Standard
	Cr	53	1423.4	7.9	**1.2652**	0.132	10.4	ug/L	328	Standard
	Mn	55	579693.3	1.0	**88.4369**	1.228	1.4	ug/L	823	Standard
	Co	59	3503.7	1.5	**0.4627**	0.007	1.6	ug/L	252	Standard
	Ni	60	3465.4	2.9	**1.7179**	0.054	3.1	ug/L	192	Standard
	Cu	65	1070.0	1.6	**0.4670**	0.008	1.6	ug/L	141	Standard
	Zn	66	3004.0	2.3	**1.8876**	0.071	3.8	ug/L	491	Standard
>	**Ge**	72	519952.2	0.4				ug/L	515087	Standard
	As	75	346.7	9.3	**0.3458**	0.028	8.2	ug/L	20	Standard
	Se	82	27.6	23.7	**0.1179**	0.056	47.6	ug/L	15	Standard
	Se-1	77	61.3	6.2	**0.0875**	0.047	53.2	ug/L	48	Standard
>	**Ga**	71	168.3	22.5				mg/L	90	Standard
	Rb	85	8202.2	4.2				ug/L	23	Standard
	Y	89	450084.2	2.9				ug/L	448510	Standard
>	**Rh**	103	20.0	50.0				ug/L	15	Standard
	Mo	98	804.4	3.7	**0.2777**	0.016	5.8	ug/L	11	Standard
	Ag	107	104.7	4.0	**-0.0007**	0.001	91.7	ug/L	134	Standard
	Cd	111	177.3	6.3	**0.0797**	0.005	6.4	mg/L	7	Standard
	Cd	114	412.5	12.7	**0.0743**	0.011	14.2	ug/L	16	Standard
>	**In**	115	613204.5	2.4				ug/L	611117	Standard
	Sn	118	838.4	3.1	**0.0394**	0.004	11.4	ug/L	518	Standard
	Sb	123	255.2	2.7	**0.0411**	0.003	6.6	ug/L	127	Standard
	Ba	135	46232.8	2.2	**22.2974**	0.510	2.3	ug/L	45	Standard
	Ce	140	10413.6	4.0				ug/L	298	Standard
>	**Tb**	159	959024.8	2.1				ug/L	977942	Standard
	Ho	165	235.0	14.0				ug/L	12	Standard
	Tl	203	65.0	35.5	**0.0028**	0.003	94.5	ug/L	31	Standard
	Tl	205	33.3	45.8	**0.0022**	0.002	93.6	ug/L	20	Standard
	Pb	206	995.4	6.0	**0.1186**	0.010	8.6	ug/L	286	Standard
	Pb	207	825.7	1.1	**0.1096**	0.003	2.3	ug/L	257	Standard
	Pb	208	3331.5	0.8	**0.1144**	0.002	1.9	ug/L	956	Standard
	U	238	237.0	6.0	**0.0298**	0.002	6.6	ug/L	17	Standard
>	**Bi**	209	562917.3	0.4				ug/L	557080	Standard

Sample ID: L1606114203
Report Date/Time: Tuesday, June 28, 2016 16:46:33
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	1456.7	5.2	**1.8709**	0.119	6.4	mg/L	32	Standard
K	39	195.0	2.6	**0.5136**	0.022	4.3	mg/L	30	Standard
Ca	43	66.7	37.7	**4.7333**	3.876	81.9	mg/L	48	Standard
Fe	54	1995.7	3.8	**1.6587**	0.089	5.4	mg/L	159	Standard
Fe	57	716.7	4.0	**1.7796**	0.068	3.8	mg/L	195	Standard
> **Sc-1**	45	39785.5	1.2				mg/L	39543	Standard
Cl	35	34494.6	2.9				ug/L	27642	Standard
Kr	83	1.3	173.2				ug/L	2	Standard
Br	81	3500.4	9.9				ug/L	1647	Standard
P	31	20160.1	3.5				ug/L	15231	Standard
S	34	2286.8	4.4				ug/L	1845	Standard
Sr	88	110.0	28.4				ug/L	70	Standard
C	12	216.7	25.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	341.7	16.9				mg/L	12	Standard
Ho-1	165	235.0	14.0				mg/L	12	Standard
Er	166	243.3	11.9				mg/L	20	Standard
I	127	25506.6	11.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		100.285	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.945	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606114203
Report Date/Time: Tuesday, June 28, 2016 16:46:33
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.342
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.048
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606114205
Sample Date/Time: Tuesday, June 28, 2016 16:47:28
Number of Replicates: 3
Autosampler Position: 245
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82787.3	1.4				ug/L	81760	Standard
	Be	9	48.3	23.9	**0.0383**	0.011	29.1	ug/L	17	Standard
	Al	27	819185.4	1.5	**5.2350**	0.125	2.4	ug/L	1410	Standard
	Sc	45	40940.2	1.5				ug/L	39543	Standard
	Ti	47	480.0	5.9	**1.6939**	0.112	6.6	ug/L	39	Standard
	V	51	6800.6	1.0	**0.7356**	0.007	0.9	ug/L	1415	Standard
	Cr	52	13980.4	1.6	**0.8771**	0.035	4.0	ug/L	6946	Standard
	Cr	53	1753.4	3.0	**1.5823**	0.062	3.9	ug/L	328	Standard
	Mn	55	596277.6	1.3	**88.4306**	1.209	1.4	ug/L	823	Standard
	Co	59	3687.4	0.3	**0.4741**	0.004	0.7	ug/L	252	Standard
	Ni	60	3717.1	1.3	**1.7958**	0.025	1.4	ug/L	192	Standard
	Cu	65	1213.0	2.2	**0.5240**	0.015	2.9	ug/L	141	Standard
	Zn	66	2977.6	2.2	**1.7886**	0.052	2.9	ug/L	491	Standard
>	**Ge**	72	534849.8	0.4				ug/L	515087	Standard
	As	75	377.9	7.0	**0.3640**	0.023	6.2	ug/L	20	Standard
	Se	82	30.4	15.6	**0.1347**	0.040	29.6	ug/L	15	Standard
	Se-1	77	66.7	7.4	**0.1333**	0.063	47.3	ug/L	48	Standard
>	**Ga**	71	223.3	16.2				mg/L	90	Standard
	Rb	85	9057.7	1.9				ug/L	23	Standard
	Y	89	448330.1	1.0				ug/L	448510	Standard
>	**Rh**	103	13.3	21.7				ug/L	15	Standard
	Mo	98	830.0	4.5	**0.2872**	0.016	5.5	ug/L	11	Standard
	Ag	107	122.7	18.8	**0.0020**	0.004	175.3	ug/L	134	Standard
	Cd	111	156.2	7.4	**0.0700**	0.006	8.4	mg/L	7	Standard
	Cd	114	373.1	12.2	**0.0667**	0.009	13.8	ug/L	16	Standard
>	**In**	115	612810.6	0.7				ug/L	611117	Standard
	Sn	118	755.0	9.6	**0.0254**	0.012	48.6	ug/L	518	Standard
	Sb	123	285.8	9.7	**0.0476**	0.006	12.8	ug/L	127	Standard
	Ba	135	48731.6	0.5	**23.5144**	0.057	0.2	ug/L	45	Standard
	Ce	140	12518.5	3.0				ug/L	298	Standard
>	**Tb**	159	979236.7	1.3				ug/L	977942	Standard
	Ho	165	273.3	12.2				ug/L	12	Standard
	Tl	203	95.3	25.1	**0.0062**	0.003	47.7	ug/L	31	Standard
	Tl	205	61.7	4.7	**0.0059**	0.000	5.9	ug/L	20	Standard
	Pb	206	1176.7	3.8	**0.1481**	0.011	7.4	ug/L	286	Standard
	Pb	207	985.7	2.0	**0.1385**	0.005	4.0	ug/L	257	Standard
	Pb	208	4033.5	3.6	**0.1475**	0.011	7.5	ug/L	956	Standard
	U	238	279.0	8.7	**0.0349**	0.004	10.9	ug/L	17	Standard
>	**Bi**	209	573358.0	1.7				ug/L	557080	Standard

Sample ID: L1606114205
Report Date/Time: Tuesday, June 28, 2016 16:49:45
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.1692**	17.605	173.1	mg/L	0	Standard
Mg	24	1471.7	7.4	**1.8343**	0.119	6.5	mg/L	32	Standard
K	39	181.7	21.4	**0.4516**	0.129	28.5	mg/L	30	Standard
Ca	43	60.0	36.3	**3.4344**	3.118	90.8	mg/L	48	Standard
Fe	54	2300.2	8.7	**1.8724**	0.148	7.9	mg/L	159	Standard
Fe	57	716.7	8.3	**1.7108**	0.158	9.3	mg/L	195	Standard
> **Sc-1**	45	40940.2	1.5				mg/L	39543	Standard
Cl	35	34657.6	1.8				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	3160.3	4.1				ug/L	1647	Standard
P	31	20585.6	2.1				ug/L	15231	Standard
S	34	2273.5	1.5				ug/L	1845	Standard
Sr	88	120.0	11.0				ug/L	70	Standard
C	12	246.7	23.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	334.1	26.2				mg/L	12	Standard
Ho-1	165	273.3	12.2				mg/L	12	Standard
Er	166	263.3	24.4				mg/L	20	Standard
I	127	21154.8	6.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.256	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.837	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606114205
Report Date/Time: Tuesday, June 28, 2016 16:49:45
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.277
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.922
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115102
Sample Date/Time: Tuesday, June 28, 2016 16:50:39
Number of Replicates: 3
Autosampler Position: 246
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83962.1	2.2				ug/L	81760	Standard
	Be	9	23.3	32.7	**0.0132**	0.008	57.7	ug/L	17	Standard
	Al	27	14475219.3	2.2	**91.4576**	2.716	3.0	ug/L	1410	Standard
	Sc	45	40821.6	0.4				ug/L	39543	Standard
	Ti	47	413.3	6.2	**1.5180**	0.138	9.1	ug/L	39	Standard
	V	51	16629.8	2.6	**2.2151**	0.113	5.1	ug/L	1415	Standard
	Cr	52	10947.0	2.1	**0.5345**	0.067	12.6	ug/L	6946	Standard
	Cr	53	3965.5	4.7	**4.2265**	0.126	3.0	ug/L	328	Standard
	Mn	55	185468.7	3.9	**28.8128**	1.697	5.9	ug/L	823	Standard
	Co	59	997.7	2.0	**0.1124**	0.006	4.9	ug/L	252	Standard
	Ni	60	1478.4	4.6	**0.6898**	0.053	7.7	ug/L	192	Standard
	Cu	65	1958.8	1.9	**0.9525**	0.040	4.2	ug/L	141	Standard
	Zn	66	8106.5	2.8	**6.6572**	0.319	4.8	ug/L	491	Standard
>	**Ge**	72	509236.4	2.2				ug/L	515087	Standard
	As	75	962.9	2.3	**0.9070**	0.035	3.9	ug/L	20	Standard
	Se	82	228.1	4.7	**1.8861**	0.120	6.4	ug/L	15	Standard
	Se-1	77	223.7	5.0	**2.2872**	0.212	9.3	ug/L	48	Standard
>	**Ga**	71	265.0	37.9				mg/L	90	Standard
	Rb	85	21039.6	2.1				ug/L	23	Standard
	Y	89	444144.2	0.6				ug/L	448510	Standard
>	**Rh**	103	50.0	20.0				ug/L	15	Standard
	Mo	98	3394.4	4.1	**1.2339**	0.051	4.2	ug/L	11	Standard
	Ag	107	107.3	13.1	**-0.0003**	0.002	751.8	ug/L	134	Standard
	Cd	111	62.6	4.9	**0.0263**	0.001	5.7	mg/L	7	Standard
	Cd	114	195.9	21.1	**0.0327**	0.008	24.5	ug/L	16	Standard
>	**In**	115	612517.0	0.3				ug/L	611117	Standard
	Sn	118	836.7	5.7	**0.0392**	0.008	20.1	ug/L	518	Standard
	Sb	123	547.1	4.7	**0.1028**	0.005	5.0	ug/L	127	Standard
	Ba	135	71666.8	2.3	**34.6081**	0.699	2.0	ug/L	45	Standard
	Ce	140	3362.0	2.6				ug/L	298	Standard
>	**Tb**	159	988372.6	0.5				ug/L	977942	Standard
	Ho	165	63.3	75.9				ug/L	12	Standard
	Tl	203	105.7	10.4	**0.0078**	0.001	16.0	ug/L	31	Standard
	Tl	205	108.3	39.8	**0.0125**	0.006	45.9	ug/L	20	Standard
	Pb	206	1800.1	2.5	**0.2697**	0.008	3.1	ug/L	286	Standard
	Pb	207	1518.4	2.6	**0.2536**	0.007	2.9	ug/L	257	Standard
	Pb	208	6212.5	1.8	**0.2692**	0.011	4.0	ug/L	956	Standard
	U	238	10075.0	2.3	**1.4043**	0.047	3.4	ug/L	17	Standard
>	**Bi**	209	556782.3	1.6				ug/L	557080	Standard

Sample ID: L1606115102
Report Date/Time: Tuesday, June 28, 2016 16:52:56
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**20.7586**	17.973	86.6	mg/L	0	Standard
Mg	24	36599.2	3.6	**47.1388**	1.781	3.8	mg/L	32	Standard
K	39	311.7	20.3	**0.8782**	0.204	23.2	mg/L	30	Standard
Ca	43	200.0	10.0	**24.0730**	2.845	11.8	mg/L	48	Standard
Fe	54	272.4	27.9	**0.1018**	0.067	65.8	mg/L	159	Standard
Fe	57	400.0	17.2	**0.6830**	0.221	32.3	mg/L	195	Standard
Sc-1	45	40821.6	0.4				mg/L	39543	Standard
Cl	35	37764.4	4.8				ug/L	27642	Standard
Kr	83	2.3	49.5				ug/L	2	Standard
Br	81	41179.3	5.9				ug/L	1647	Standard
P	31	20392.0	1.6				ug/L	15231	Standard
S	34	2350.2	4.9				ug/L	1845	Standard
Sr	88	246.7	3.1				ug/L	70	Standard
C	12	240.0	16.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	93.2	27.3				mg/L	12	Standard
Ho-1	165	63.3	75.9				mg/L	12	Standard
Er	166	73.3	7.9				mg/L	20	Standard
I	127	267945.4	8.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		102.693	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		98.864	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606115102
Report Date/Time: Tuesday, June 28, 2016 16:52:56
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.229
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.946
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115105
Sample Date/Time: Tuesday, June 28, 2016 16:53:51
Number of Replicates: 3
Autosampler Position: 247
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81681.4	3.1				ug/L	81760	Standard
	Be	9	16.7	45.8	**0.0069**	0.007	104.2	ug/L	17	Standard
	Al	27	18070414.8	1.1	**117.3803**	2.444	2.1	ug/L	1410	Standard
	Sc	45	40721.3	2.6				ug/L	39543	Standard
	Ti	47	244.7	8.0	**0.8457**	0.069	8.1	ug/L	39	Standard
	V	51	17618.4	3.0	**2.3985**	0.046	1.9	ug/L	1415	Standard
	Cr	52	10888.2	1.3	**0.5506**	0.002	0.4	ug/L	6946	Standard
	Cr	53	4290.6	7.0	**4.6855**	0.402	8.6	ug/L	328	Standard
	Mn	55	234137.7	1.0	**36.9743**	0.056	0.2	ug/L	823	Standard
	Co	59	907.0	1.1	**0.1013**	0.001	0.9	ug/L	252	Standard
	Ni	60	1595.4	2.1	**0.7658**	0.017	2.3	ug/L	192	Standard
	Cu	65	1678.8	2.6	**0.8172**	0.017	2.1	ug/L	141	Standard
	Zn	66	5525.3	1.2	**4.3540**	0.048	1.1	ug/L	491	Standard
>	**Ge**	72	501124.8	1.2				ug/L	515087	Standard
	As	75	1072.7	4.1	**1.0212**	0.046	4.5	ug/L	20	Standard
	Se	82	264.0	5.9	**2.2378**	0.139	6.2	ug/L	15	Standard
	Se-1	77	284.0	16.7	**3.1586**	0.667	21.1	ug/L	48	Standard
>	**Ga**	71	116.7	13.8				mg/L	90	Standard
	Rb	85	15923.4	4.0				ug/L	23	Standard
	Y	89	444308.5	1.8				ug/L	448510	Standard
>	**Rh**	103	68.3	40.3				ug/L	15	Standard
	Mo	98	3823.6	1.7	**1.4182**	0.006	0.4	ug/L	11	Standard
	Ag	107	76.3	4.0	**-0.0047**	0.001	10.8	ug/L	134	Standard
	Cd	111	40.8	13.4	**0.0165**	0.003	15.5	mg/L	7	Standard
	Cd	114	111.3	16.3	**0.0169**	0.004	22.4	ug/L	16	Standard
>	**In**	115	601437.2	1.6				ug/L	611117	Standard
	Sn	118	1256.7	47.3	**0.1130**	0.098	86.7	ug/L	518	Standard
	Sb	123	502.0	6.8	**0.0953**	0.009	9.3	ug/L	127	Standard
	Ba	135	85456.0	1.5	**42.0358**	0.440	1.0	ug/L	45	Standard
	Ce	140	1221.7	20.4				ug/L	298	Standard
>	**Tb**	159	951914.7	0.8				ug/L	977942	Standard
	Ho	165	38.3	39.8				ug/L	12	Standard
	Tl	203	109.0	4.0	**0.0087**	0.001	7.8	ug/L	31	Standard
	Tl	205	100.0	13.2	**0.0120**	0.002	15.9	ug/L	20	Standard
	Pb	206	942.7	3.4	**0.1187**	0.004	3.7	ug/L	286	Standard
	Pb	207	751.4	3.5	**0.1031**	0.007	7.3	ug/L	257	Standard
	Pb	208	3320.1	1.9	**0.1235**	0.005	4.4	ug/L	956	Standard
	U	238	12716.0	1.7	**1.8521**	0.048	2.6	ug/L	17	Standard
>	**Bi**	209	533061.4	1.1				ug/L	557080	Standard

Sample ID: L1606115105
Report Date/Time: Tuesday, June 28, 2016 16:56:08
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.4978**	18.174	173.1	mg/L	0	Standard
Mg	24	45127.2	3.7	**58.2718**	1.297	2.2	mg/L	32	Standard
K	39	268.3	17.6	**0.7391**	0.156	21.1	mg/L	30	Standard
Ca	43	230.0	5.8	**28.6229**	2.753	9.6	mg/L	48	Standard
Fe	54	227.2	16.5	**0.0631**	0.037	58.2	mg/L	159	Standard
Fe	57	321.7	8.6	**0.4294**	0.079	18.3	mg/L	195	Standard
> **Sc-1**	45	40721.3	2.6				mg/L	39543	Standard
Cl	35	40535.5	2.1				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	42472.8	3.2				ug/L	1647	Standard
P	31	21144.7	2.2				ug/L	15231	Standard
S	34	2520.2	4.3				ug/L	1845	Standard
Sr	88	268.3	23.3				ug/L	70	Standard
C	12	270.0	30.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	48.6	21.1				mg/L	12	Standard
Ho-1	165	38.3	39.8				mg/L	12	Standard
Er	166	30.0	33.3				mg/L	20	Standard
I	127	357293.3	4.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		99.904	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.289	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115105
Report Date/Time: Tuesday, June 28, 2016 16:56:08
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.416
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.688
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606115105
Report Date/Time: Tuesday, June 28, 2016 16:56:08
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115401 WG573751-03
Sample Date/Time: Tuesday, June 28, 2016 16:57:02
Number of Replicates: 3
Autosampler Position: 248
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100430.3	2.8				ug/L	81760	Standard
	Be	9	28.3	56.7	**0.0134**	0.013	96.3	ug/L	17	Standard
	Al	27	81335563.7	1.7	**429.6767**	8.046	1.9	ug/L	1410	Standard
	Sc	45	36054.5	1.3				ug/L	39543	Standard
	Ti	47	276.7	4.4	**1.0861**	0.041	3.8	ug/L	39	Standard
	V	51	56034.7	2.7	**8.8625**	0.243	2.7	ug/L	1415	Standard
	Cr	52	12377.7	0.7	**0.9433**	0.019	2.0	ug/L	6946	Standard
	Cr	53	23940.6	3.3	**30.3443**	1.215	4.0	ug/L	328	Standard
	Mn	55	741944.6	0.9	**128.6152**	2.806	2.2	ug/L	823	Standard
	Co	59	3502.1	2.3	**0.5295**	0.021	4.0	ug/L	252	Standard
	Ni	60	10135.4	1.6	**5.9492**	0.102	1.7	ug/L	192	Standard
	Cu	65	4993.5	1.7	**2.8661**	0.085	3.0	ug/L	141	Standard
	Zn	66	5821.5	1.7	**5.1486**	0.184	3.6	ug/L	491	Standard
>	**Ge**	72	457910.9	1.5				ug/L	515087	Standard
	As	75	1957.2	6.1	**1.9997**	0.145	7.2	ug/L	20	Standard
	Se	82	3071.1	3.5	**29.9014**	1.486	5.0	ug/L	15	Standard
	Se-1	77	4244.3	1.8	**62.6742**	2.070	3.3	ug/L	48	Standard
>	**Ga**	71	138.3	10.4				mg/L	90	Standard
	Rb	85	9993.3	0.6				ug/L	23	Standard
	Y	89	397349.5	0.4				ug/L	448510	Standard
>	**Rh**	103	1400.1	6.9				ug/L	15	Standard
	Mo	98	28731.5	0.6	**12.2349**	0.110	0.9	ug/L	11	Standard
	Ag	107	395.7	2.6	**0.0517**	0.002	4.3	ug/L	134	Standard
	Cd	111	92.0	15.6	**0.0467**	0.008	17.2	mg/L	7	Standard
	Cd	114	319.4	5.6	**0.0659**	0.004	5.6	ug/L	16	Standard
>	**In**	115	530125.1	0.7				ug/L	611117	Standard
	Sn	118	1140.0	5.3	**0.1203**	0.011	8.8	ug/L	518	Standard
	Sb	123	532.5	6.0	**0.1172**	0.007	5.9	ug/L	127	Standard
	Ba	135	16736.6	1.8	**9.3206**	0.182	2.0	ug/L	45	Standard
	Ce	140	1620.1	9.0				ug/L	298	Standard
>	**Tb**	159	890828.2	2.0				ug/L	977942	Standard
	Ho	165	38.3	27.2				ug/L	12	Standard
	Tl	203	267.7	7.7	**0.0327**	0.003	10.1	ug/L	31	Standard
	Tl	205	231.7	6.6	**0.0350**	0.002	6.9	ug/L	20	Standard
	Pb	206	883.7	4.8	**0.1290**	0.007	5.4	ug/L	286	Standard
	Pb	207	738.7	4.5	**0.1206**	0.010	8.0	ug/L	257	Standard
	Pb	208	2811.4	0.6	**0.1152**	0.004	3.5	ug/L	956	Standard
	U	238	63722.1	1.7	**10.4767**	0.095	0.9	ug/L	17	Standard
>	**Bi**	209	472777.8	1.7				ug/L	557080	Standard

Sample ID: L1606115401 WG573751-03
Report Date/Time: Tuesday, June 28, 2016 16:59:19
Page 1

Approved: June 29, 2016

Na	23	15.0	33.3	**105.9031**	36.283	34.3	mg/L	0	Standard
Mg	24	356514.1	0.8	**520.5310**	10.800	2.1	mg/L	32	Standard
K	39	430.0	2.3	**1.4525**	0.056	3.9	mg/L	30	Standard
Ca	43	1840.1	5.6	**301.0738**	13.488	4.5	mg/L	48	Standard
Fe	54	188.6	12.1	**0.0500**	0.021	41.8	mg/L	159	Standard
Fe	57	1585.1	2.5	**5.2457**	0.200	3.8	mg/L	195	Standard
> **Sc-1**	45	36054.5	1.3				mg/L	39543	Standard
Cl	35	90598.1	1.3				ug/L	27642	Standard
Kr	83	3.0	66.7				ug/L	2	Standard
Br	81	42312.4	3.9				ug/L	1647	Standard
P	31	19347.4	2.8				ug/L	15231	Standard
S	34	2421.9	2.4				ug/L	1845	Standard
Sr	88	6231.3	0.9				ug/L	70	Standard
C	12	243.3	12.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	37.9	25.6				mg/L	12	Standard
Ho-1	165	38.3	27.2				mg/L	12	Standard
Er	166	43.3	96.1				mg/L	20	Standard
I	127	118064.4	6.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		122.835	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		88.900	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115401 WG573751-03
Report Date/Time: Tuesday, June 28, 2016 16:59:19
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	86.747
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	84.867
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606115401 WG573751-03
Report Date/Time: Tuesday, June 28, 2016 16:59:19
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115402S WG573751-09
Sample Date/Time: Tuesday, June 28, 2016 17:00:14
Number of Replicates: 3
Autosampler Position: 249
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	111794.4	2.2				ug/L	81760	Standard
	Be	9	61074.6	3.8	**44.9573**	0.853	1.9	ug/L	17	Standard
	Al	27	84589452.8	4.6	**401.1886**	9.792	2.4	ug/L	1410	Standard
	Sc	45	39148.9	4.1				ug/L	39543	Standard
	Ti	47	272.0	24.8	**0.9934**	0.318	32.0	ug/L	39	Standard
	V	51	406216.4	3.4	**61.4172**	0.609	1.0	ug/L	1415	Standard
	Cr	52	347340.1	3.9	**51.5532**	0.830	1.6	ug/L	6946	Standard
	Cr	53	63548.0	2.3	**76.0057**	0.198	0.3	ug/L	328	Standard
	Mn	55	986761.6	2.9	**160.3601**	1.162	0.7	ug/L	823	Standard
	Co	59	344414.5	2.6	**51.6648**	0.567	1.1	ug/L	252	Standard
	Ni	60	96461.7	3.6	**53.9005**	0.670	1.2	ug/L	192	Standard
	Cu	65	88941.7	2.7	**49.2712**	0.171	0.3	ug/L	141	Standard
	Zn	66	55366.0	3.1	**52.4384**	0.443	0.8	ug/L	491	Standard
>	**Ge**	72	488431.6	2.5				ug/L	515087	Standard
	As	75	59145.7	4.0	**55.5035**	0.877	1.6	ug/L	20	Standard
	Se	82	9257.9	4.7	**84.6512**	1.891	2.2	ug/L	15	Standard
	Se-1	77	8241.9	2.0	**114.6628**	0.599	0.5	ug/L	48	Standard
>	**Ga**	71	135.0	3.7				mg/L	90	Standard
	Rb	85	11232.5	8.2				ug/L	23	Standard
	Y	89	424780.5	4.2				ug/L	448510	Standard
>	**Rh**	103	1410.1	5.3				ug/L	15	Standard
	Mo	98	28554.4	4.0	**11.2353**	0.049	0.4	ug/L	11	Standard
	Ag	107	298500.0	4.4	**47.3531**	0.325	0.7	ug/L	134	Standard
	Cd	111	96838.7	4.3	**48.2572**	0.264	0.5	mg/L	7	Standard
	Cd	114	236461.6	3.9	**48.4818**	0.307	0.6	ug/L	16	Standard
>	**In**	115	573601.2	3.7				ug/L	611117	Standard
	Sn	118	926.7	16.3	**0.0649**	0.025	39.2	ug/L	518	Standard
	Sb	123	232072.7	4.2	**52.3852**	0.509	1.0	ug/L	127	Standard
	Ba	135	113861.5	4.7	**58.7202**	0.562	1.0	ug/L	45	Standard
	Ce	140	1120.0	4.0				ug/L	298	Standard
>	**Tb**	159	960238.4	4.5				ug/L	977942	Standard
	Ho	165	33.3	75.5				ug/L	12	Standard
	Tl	203	395828.3	3.1	**51.4711**	0.762	1.5	ug/L	31	Standard
	Tl	205	346560.1	3.5	**51.8878**	0.159	0.3	ug/L	20	Standard
	Pb	206	263890.3	2.7	**53.4438**	0.979	1.8	ug/L	286	Standard
	Pb	207	230261.5	2.5	**51.7030**	0.928	1.8	ug/L	257	Standard
	Pb	208	887552.6	2.9	**51.5030**	0.554	1.1	ug/L	956	Standard
	U	238	419873.0	2.9	**64.2072**	0.642	1.0	ug/L	17	Standard
>	**Bi**	209	508536.8	3.8				ug/L	557080	Standard

Sample ID: L1606115402S WG573751-09
Report Date/Time: Tuesday, June 28, 2016 17:02:31
Page 1

Approved: June 29, 2016

Na	23	21.7	35.3	**139.6053**	45.304	32.5	mg/L	0	Standard
Mg	24	357824.7	2.8	**481.2852**	10.577	2.2	mg/L	32	Standard
K	39	491.7	6.5	**1.5399**	0.155	10.1	mg/L	30	Standard
Ca	43	1775.1	1.7	**267.3999**	15.688	5.9	mg/L	48	Standard
Fe	54	171.9	15.3	**0.0199**	0.020	102.8	mg/L	159	Standard
Fe	57	1550.1	0.6	**4.6666**	0.208	4.5	mg/L	195	Standard
Sc-1	45	39148.9	4.1				mg/L	39543	Standard
Cl	35	91435.3	3.5				ug/L	27642	Standard
Kr	83	2.7	21.7				ug/L	2	Standard
Br	81	45418.1	6.1				ug/L	1647	Standard
P	31	20330.3	5.5				ug/L	15231	Standard
S	34	2288.5	3.3				ug/L	1845	Standard
Sr	88	6416.4	6.4				ug/L	70	Standard
C	12	206.7	37.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	21.9	50.9				mg/L	12	Standard
Ho-1	165	33.3	75.5				mg/L	12	Standard
Er	166	30.0	57.7				mg/L	20	Standard
I	127	49499.1	2.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		136.735	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.825	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115402S WG573751-09
Report Date/Time: Tuesday, June 28, 2016 17:02:31
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	93.861
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	91.286
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606115402S WG573751-09
Report Date/Time: Tuesday, June 28, 2016 17:02:31
Page 3

Approved: June 29, 2016

Se-1 77 Upper, S, EEE	Se-1	77

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115403SD WG573751-10
Sample Date/Time: Tuesday, June 28, 2016 17:03:25
Number of Replicates: 3
Autosampler Position: 250
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	118994.0	2.6				ug/L	81760	Standard
	Be	9	64654.3	0.9	**44.7357**	0.820	1.8	ug/L	17	Standard
	Al	27	87741403.7	3.4	**391.0655**	6.582	1.7	ug/L	1410	Standard
	Sc	45	41172.5	1.6				ug/L	39543	Standard
	Ti	47	259.3	3.7	**0.9238**	0.057	6.2	ug/L	39	Standard
	V	51	425385.5	3.2	**63.6566**	1.295	2.0	ug/L	1415	Standard
	Cr	52	362069.3	1.9	**53.2267**	0.604	1.1	ug/L	6946	Standard
	Cr	53	64441.8	2.0	**76.2717**	0.872	1.1	ug/L	328	Standard
	Mn	55	1073606.3	3.1	**172.6721**	4.096	2.4	ug/L	823	Standard
	Co	59	369825.3	3.2	**54.8942**	1.128	2.1	ug/L	252	Standard
	Ni	60	100548.2	1.5	**55.6131**	0.149	0.3	ug/L	192	Standard
	Cu	65	92359.0	1.3	**50.6385**	0.319	0.6	ug/L	141	Standard
	Zn	66	57865.3	2.6	**54.2651**	0.864	1.6	ug/L	491	Standard
>	**Ge**	72	493572.0	1.8				ug/L	515087	Standard
	As	75	61885.0	3.3	**57.4777**	1.381	2.4	ug/L	20	Standard
	Se	82	9664.4	3.7	**87.4717**	2.250	2.6	ug/L	15	Standard
	Se-1	77	8366.3	3.5	**115.1653**	2.897	2.5	ug/L	48	Standard
>	**Ga**	71	118.3	21.7				mg/L	90	Standard
	Rb	85	11170.8	6.4				ug/L	23	Standard
	Y	89	447220.0	3.1				ug/L	448510	Standard
>	**Rh**	103	1640.1	8.8				ug/L	15	Standard
	Mo	98	29726.0	4.1	**11.0857**	0.160	1.4	ug/L	11	Standard
	Ag	107	312572.9	2.2	**47.0220**	0.762	1.6	ug/L	134	Standard
	Cd	111	102356.1	3.3	**48.3584**	0.794	1.6	mg/L	7	Standard
	Cd	114	246212.9	4.6	**47.8408**	0.884	1.8	ug/L	16	Standard
>	**In**	115	605206.5	3.7				ug/L	611117	Standard
	Sn	118	1233.4	10.6	**0.1085**	0.018	16.3	ug/L	518	Standard
	Sb	123	240651.4	1.9	**51.5127**	0.935	1.8	ug/L	127	Standard
	Ba	135	117995.0	3.4	**57.6913**	0.169	0.3	ug/L	45	Standard
	Ce	140	1415.1	2.3				ug/L	298	Standard
>	**Tb**	159	994298.0	3.0				ug/L	977942	Standard
	Ho	165	46.7	12.4				ug/L	12	Standard
	Tl	203	412661.0	1.3	**51.8933**	0.197	0.4	ug/L	31	Standard
	Tl	205	366877.2	3.6	**53.1179**	1.107	2.1	ug/L	20	Standard
	Pb	206	275320.1	2.3	**53.9118**	0.413	0.8	ug/L	286	Standard
	Pb	207	238830.1	2.6	**51.8469**	0.552	1.1	ug/L	257	Standard
	Pb	208	936516.1	2.9	**52.5472**	0.756	1.4	ug/L	956	Standard
	U	238	442147.9	3.9	**65.3698**	1.598	2.4	ug/L	17	Standard
>	**Bi**	209	525736.2	1.5				ug/L	557080	Standard

Sample ID: L1606115403SD WG573751-10
Report Date/Time: Tuesday, June 28, 2016 17:05:42
Page 1

Approved: June 29, 2016

Na	23	31.7	48.2	**194.7966**	93.559	48.0	mg/L	0	Standard
Mg	24	381865.2	2.0	**488.2175**	12.226	2.5	mg/L	32	Standard
K	39	475.0	11.9	**1.3985**	0.159	11.4	mg/L	30	Standard
Ca	43	1828.4	7.5	**261.2707**	17.620	6.7	mg/L	48	Standard
Fe	54	210.4	5.2	**0.0459**	0.012	26.9	mg/L	159	Standard
Fe	57	1693.4	5.0	**4.8640**	0.189	3.9	mg/L	195	Standard
> **Sc-1**	45	41172.5	1.6				mg/L	39543	Standard
Cl	35	96785.1	3.5				ug/L	27642	Standard
Kr	83	2.7	78.1				ug/L	2	Standard
Br	81	48665.0	7.0				ug/L	1647	Standard
P	31	21074.7	3.4				ug/L	15231	Standard
S	34	2411.9	0.6				ug/L	1845	Standard
Sr	88	7001.6	9.4				ug/L	70	Standard
C	12	253.3	28.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	54.1	45.2				mg/L	12	Standard
Ho-1	165	46.7	12.4				mg/L	12	Standard
Er	166	53.3	28.6				mg/L	20	Standard
I	127	96813.1	12.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		145.540	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.823	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115403SD WG573751-10
Report Date/Time: Tuesday, June 28, 2016 17:05:42
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	99.033
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	94.373
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606115403SD WG573751-10
Report Date/Time: Tuesday, June 28, 2016 17:05:42
Page 3

Approved: June 29, 2016

Se-1 77 Upper, S, EEE	Se-1	77

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115404
Sample Date/Time: Tuesday, June 28, 2016 17:06:37
Number of Replicates: 3
Autosampler Position: 251
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100804.3	1.4				ug/L	81760	Standard
	Be	9	43.3	6.7	**0.0257**	0.002	9.1	ug/L	17	Standard
	Al	27	95134.9	5.2	**0.4862**	0.026	5.4	ug/L	1410	Standard
	Sc	45	38767.9	2.5				ug/L	39543	Standard
	Ti	47	56.0	17.6	**0.0568**	0.042	74.0	ug/L	39	Standard
	V	51	2653.2	7.2	**0.1688**	0.023	13.6	ug/L	1415	Standard
	Cr	52	16846.0	0.3	**1.3242**	0.040	3.0	ug/L	6946	Standard
	Cr	53	2448.5	10.3	**2.4055**	0.318	13.2	ug/L	328	Standard
	Mn	55	6416.4	0.3	**0.8254**	0.012	1.5	ug/L	823	Standard
	Co	59	385.0	5.1	**0.0226**	0.003	13.2	ug/L	252	Standard
	Ni	60	1595.4	9.2	**0.7282**	0.071	9.8	ug/L	192	Standard
	Cu	65	1383.7	5.1	**0.6260**	0.046	7.3	ug/L	141	Standard
	Zn	66	3329.0	1.8	**2.1614**	0.092	4.3	ug/L	491	Standard
>	**Ge**	72	523540.2	1.3				ug/L	515087	Standard
	As	75	21.6	288.4	**0.0594**	0.054	91.4	ug/L	20	Standard
	Se	82	48.4	30.2	**0.2956**	0.131	44.3	ug/L	15	Standard
	Se-1	77	118.7	6.2	**0.8319**	0.117	14.0	ug/L	48	Standard
>	**Ga**	71	35.0	51.5				mg/L	90	Standard
	Rb	85	285.0	10.7				ug/L	23	Standard
	Y	89	449294.4	0.5				ug/L	448510	Standard
>	**Rh**	103	35.0	14.3				ug/L	15	Standard
	Mo	98	461.6	2.9	**0.1485**	0.005	3.0	ug/L	11	Standard
	Ag	107	143.3	7.2	**0.0047**	0.002	32.8	ug/L	134	Standard
	Cd	111	43.9	7.3	**0.0172**	0.002	8.8	mg/L	7	Standard
	Cd	114	151.5	5.9	**0.0237**	0.002	7.4	ug/L	16	Standard
>	**In**	115	622732.4	0.3				ug/L	611117	Standard
	Sn	118	3920.5	2.1	**0.5488**	0.015	2.8	ug/L	518	Standard
	Sb	123	270.3	27.2	**0.0434**	0.015	35.5	ug/L	127	Standard
	Ba	135	1327.1	1.7	**0.6057**	0.011	1.9	ug/L	45	Standard
	Ce	140	168.3	16.4				ug/L	298	Standard
>	**Tb**	159	980536.6	0.9				ug/L	977942	Standard
	Ho	165	21.7	13.3				ug/L	12	Standard
	Tl	203	266.3	8.1	**0.0259**	0.003	10.5	ug/L	31	Standard
	Tl	205	201.7	5.7	**0.0244**	0.001	5.0	ug/L	20	Standard
	Pb	206	459.0	5.5	**0.0188**	0.005	29.2	ug/L	286	Standard
	Pb	207	389.7	6.3	**0.0195**	0.006	29.5	ug/L	257	Standard
	Pb	208	1625.7	6.2	**0.0231**	0.006	26.8	ug/L	956	Standard
	U	238	121.7	7.4	**0.0135**	0.001	9.9	ug/L	17	Standard
>	**Bi**	209	574701.5	1.2				ug/L	557080	Standard

Sample ID: L1606115404
Report Date/Time: Tuesday, June 28, 2016 17:08:54
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**32.7163**	32.227	98.5	mg/L	0	Standard
Mg	24	330.0	22.9	**0.3898**	0.094	24.2	mg/L	32	Standard
K	39	33.3	43.3	**-0.0268**	0.048	179.7	mg/L	30	Standard
Ca	43	33.3	43.3	**-0.2011**	2.146	1067.4	mg/L	48	Standard
Fe	54	141.8	17.1	**-0.0059**	0.025	416.6	mg/L	159	Standard
Fe	57	216.7	23.3	**0.1207**	0.166	137.9	mg/L	195	Standard
> **Sc-1**	45	38767.9	2.5				mg/L	39543	Standard
Cl	35	43132.0	0.5				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	3107.0	8.7				ug/L	1647	Standard
P	31	17094.6	3.8				ug/L	15231	Standard
S	34	2171.8	4.7				ug/L	1845	Standard
Sr	88	71.7	31.5				ug/L	70	Standard
C	12	180.0	44.1				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	6.3	86.6				mg/L	12	Standard
Ho-1	165	21.7	13.3				mg/L	12	Standard
Er	166	6.7	86.6				mg/L	20	Standard
I	127	6683.3	38.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		123.293	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		101.641	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606115404
Report Date/Time: Tuesday, June 28, 2016 17:08:54
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	101.901
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	103.163
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 17:09:50
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	98106.9	3.9				ug/L	81760	Standard
	Be	9	60781.9	5.8	**50.9762**	1.224	2.4	ug/L	17	Standard
	Al	27	8271215.1	2.8	**44.7180**	0.568	1.3	ug/L	1410	Standard
	Sc	45	39052.0	5.6				ug/L	39543	Standard
	Ti	47	24646.4	4.4	**101.4481**	2.619	2.6	ug/L	39	Standard
	V	51	347970.7	1.4	**51.1411**	0.534	1.0	ug/L	1415	Standard
	Cr	52	354198.6	3.4	**51.1178**	0.779	1.5	ug/L	6946	Standard
	Cr	53	44595.6	4.3	**51.7475**	1.396	2.7	ug/L	328	Standard
	Mn	55	315347.9	1.9	**49.7410**	0.694	1.4	ug/L	823	Standard
	Co	59	365725.5	2.3	**53.3530**	0.828	1.6	ug/L	252	Standard
	Ni	60	93315.9	2.3	**50.7083**	0.301	0.6	ug/L	192	Standard
	Cu	65	94072.6	2.8	**50.6778**	0.428	0.8	ug/L	141	Standard
	Zn	66	55586.5	1.6	**51.1865**	0.432	0.8	ug/L	491	Standard
>	Ge	72	502280.1	2.3				ug/L	515087	Standard
	As	75	55868.4	1.2	**51.0087**	0.746	1.5	ug/L	20	Standard
	Se	82	5921.1	2.6	**52.6309**	1.371	2.6	ug/L	15	Standard
	Se-1	77	3936.2	0.2	**52.8821**	1.332	2.5	ug/L	48	Standard
>	Ga	71	43.3	6.7				mg/L	90	Standard
	Rb	85	853.4	12.0				ug/L	23	Standard
	Y	89	445806.9	0.5				ug/L	448510	Standard
>	Rh	103	45.0	61.9				ug/L	15	Standard
	Mo	98	273081.8	2.1	**98.9265**	1.774	1.8	ug/L	11	Standard
	Ag	107	339843.9	2.2	**49.5651**	0.762	1.5	ug/L	134	Standard
	Cd	111	106758.1	1.3	**48.9107**	0.367	0.7	mg/L	7	Standard
	Cd	114	258025.9	1.5	**48.6321**	0.403	0.8	ug/L	16	Standard
>	In	115	623979.7	0.7				ug/L	611117	Standard
	Sn	118	298451.0	1.1	**49.3312**	0.224	0.5	ug/L	518	Standard
	Sb	123	238781.4	2.0	**49.5506**	0.796	1.6	ug/L	127	Standard
	Ba	135	101752.9	1.4	**48.2444**	0.388	0.8	ug/L	45	Standard
	Ce	140	83.3	6.9				ug/L	298	Standard
>	Tb	159	999812.3	3.7				ug/L	977942	Standard
	Ho	165	18.3	56.8				ug/L	12	Standard
	Tl	203	438487.8	2.0	**50.5214**	0.372	0.7	ug/L	31	Standard
	Tl	205	396911.7	0.7	**52.6785**	1.200	2.3	ug/L	20	Standard
	Pb	206	278860.8	1.5	**50.0350**	0.542	1.1	ug/L	286	Standard
	Pb	207	252734.9	2.4	**50.2714**	0.461	0.9	ug/L	257	Standard
	Pb	208	1008405.1	1.0	**51.8566**	0.775	1.5	ug/L	956	Standard
	U	238	385967.4	0.9	**52.3060**	0.849	1.6	ug/L	17	Standard
>	Bi	209	573840.1	2.4				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:12:07
Page 1

Approved: June 29, 2016

Na	23	6.7	43.3	**42.7715**	15.957	37.3	mg/L	0	Standard
Mg	24	4330.6	4.7	**5.7850**	0.241	4.2	mg/L	32	Standard
K	39	1310.1	5.5	**4.3410**	0.087	2.0	mg/L	30	Standard
Ca	43	75.0	17.6	**6.2533**	2.500	40.0	mg/L	48	Standard
Fe	54	6082.2	3.3	**5.4477**	0.350	6.4	mg/L	159	Standard
Fe	57	1690.1	5.8	**5.1531**	0.140	2.7	mg/L	195	Standard
Sc-1	45	39052.0	5.6				mg/L	39543	Standard
Cl	35	36312.2	3.9				ug/L	27642	Standard
Kr	83	2.7	57.3				ug/L	2	Standard
Br	81	2130.1	3.3				ug/L	1647	Standard
P	31	18776.6	1.2				ug/L	15231	Standard
S	34	2388.5	6.0				ug/L	1845	Standard
Sr	88	65.0	13.3				ug/L	70	Standard
C	12	133.3	33.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	12.7	43.5				mg/L	12	Standard
Ho-1	165	18.3	56.8				mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	2708.6	18.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	101.952		
Al	27	89.436		
Sc	45			
Ti	47	101.448		
V	51	102.282		
Cr	52	102.236		
Cr	53			
Mn	55	99.482		
Co	59	106.706		
Ni	60	101.417		
Cu	65	101.356		
Zn	66	102.373		
Ge	72		97.514	
As	75	102.017		
Se	82	105.262		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:12:07
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	98.926		
Ag	107	99.130		
Cd	111	97.821		
Cd	114			
In	115		102.105	
Sn	118	98.662		
Sb	123	99.101		
Ba	135	96.489		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.043		
Tl	205			
Pb	206			
Pb	207			
Pb	208	103.713		
U	238	104.612		
Bi	209		103.008	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:12:07
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 17:13:01
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92565.4	2.9				ug/L	81760	Standard
	Be	9	53.3	124.8	**0.0389**	0.062	158.6	ug/L	17	Standard
	Al	27	40870.6	146.9	**0.2266**	0.357	157.4	ug/L	1410	Standard
	Sc	45	37247.5	4.3				ug/L	39543	Standard
	Ti	47	28.7	28.0	**-0.0471**	0.033	70.8	ug/L	39	Standard
	V	51	1306.5	27.0	**-0.0143**	0.053	368.2	ug/L	1415	Standard
	Cr	52	6317.0	3.9	**-0.1289**	0.039	29.9	ug/L	6946	Standard
	Cr	53	723.4	11.4	**0.5015**	0.098	19.5	ug/L	328	Standard
	Mn	55	1084.7	54.0	**0.0228**	0.093	407.7	ug/L	823	Standard
	Co	59	338.3	37.8	**0.0179**	0.019	104.7	ug/L	252	Standard
	Ni	60	322.7	24.3	**0.0703**	0.043	61.2	ug/L	192	Standard
	Cu	65	237.7	21.2	**0.0371**	0.027	74.0	ug/L	141	Standard
	Zn	66	487.7	9.0	**-0.3725**	0.041	11.1	ug/L	491	Standard
>	**Ge**	72	503592.6	0.3				ug/L	515087	Standard
	As	75	-19.3	188.6	**0.0227**	0.033	146.0	ug/L	20	Standard
	Se	82	22.8	34.7	**0.0835**	0.070	83.8	ug/L	15	Standard
	Se-1	77	71.3	3.2	**0.2496**	0.032	12.8	ug/L	48	Standard
>	**Ga**	71	43.3	46.6				mg/L	90	Standard
	Rb	85	33.3	37.7				ug/L	23	Standard
	Y	89	434857.2	0.6				ug/L	448510	Standard
>	**Rh**	103	13.3	43.3				ug/L	15	Standard
	Mo	98	288.2	19.1	**0.0890**	0.021	23.1	ug/L	11	Standard
	Ag	107	295.0	92.5	**0.0287**	0.042	146.5	ug/L	134	Standard
	Cd	111	58.8	115.8	**0.0253**	0.033	130.1	mg/L	7	Standard
	Cd	114	179.7	150.5	**0.0306**	0.054	175.0	ug/L	16	Standard
>	**In**	115	602649.0	1.7				ug/L	611117	Standard
	Sn	118	880.0	23.1	**0.0494**	0.037	75.1	ug/L	518	Standard
	Sb	123	564.4	53.6	**0.1091**	0.067	61.4	ug/L	127	Standard
	Ba	135	113.7	104.6	**0.0312**	0.059	190.7	ug/L	45	Standard
	Ce	140	26.7	21.7				ug/L	298	Standard
>	**Tb**	159	956287.2	0.6				ug/L	977942	Standard
	Ho	165	6.7	173.2				ug/L	12	Standard
	Tl	203	302.3	107.0	**0.0307**	0.038	123.1	ug/L	31	Standard
	Tl	205	353.3	119.4	**0.0454**	0.057	124.8	ug/L	20	Standard
	Pb	206	425.3	47.1	**0.0143**	0.036	251.9	ug/L	286	Standard
	Pb	207	353.0	32.9	**0.0137**	0.023	168.1	ug/L	257	Standard
	Pb	208	1674.0	55.6	**0.0274**	0.048	175.9	ug/L	956	Standard
	U	238	298.0	132.6	**0.0381**	0.054	142.5	ug/L	17	Standard
>	**Bi**	209	561401.5	0.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:15:18
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	206.7	141.8	**0.2403**	0.424	176.4	mg/L	32	Standard
K	39	28.3	71.3	**-0.0379**	0.076	200.2	mg/L	30	Standard
Ca	43	35.0	24.7	**0.2817**	1.299	461.1	mg/L	48	Standard
Fe	54	147.7	17.2	**0.0057**	0.030	525.6	mg/L	159	Standard
Fe	57	195.0	14.3	**0.0744**	0.099	133.6	mg/L	195	Standard
> **Sc-1**	45	37247.5	4.3				mg/L	39543	Standard
Cl	35	35032.5	2.2				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1893.5	11.2				ug/L	1647	Standard
P	31	18479.6	2.3				ug/L	15231	Standard
S	34	2273.5	3.9				ug/L	1845	Standard
Sr	88	81.7	3.5				ug/L	70	Standard
C	12	186.7	15.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	9.5	186.3				mg/L	12	Standard
Ho-1	165	6.7	173.2				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	2043.5	19.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.768	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:15:18
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.614
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.776
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:15:18
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115501
Sample Date/Time: Tuesday, June 28, 2016 17:16:14
Number of Replicates: 3
Autosampler Position: 252
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92590.4	1.6				ug/L	81760	Standard
	Be	9	26.7	47.2	**0.0141**	0.012	82.4	ug/L	17	Standard
	Al	27	22123057.0	1.6	**126.7281**	1.652	1.3	ug/L	1410	Standard
	Sc	45	39028.5	1.1				ug/L	39543	Standard
	Ti	47	160.7	8.5	**0.5119**	0.051	9.9	ug/L	39	Standard
	V	51	4388.8	2.7	**0.4554**	0.016	3.5	ug/L	1415	Standard
	Cr	52	10159.7	1.1	**0.4737**	0.053	11.2	ug/L	6946	Standard
	Cr	53	1711.8	8.9	**1.7064**	0.233	13.7	ug/L	328	Standard
	Mn	55	4816073.5	0.9	**779.0277**	27.654	3.5	ug/L	823	Standard
	Co	59	2714.6	0.9	**0.3738**	0.009	2.5	ug/L	252	Standard
	Ni	60	3772.1	1.1	**1.9951**	0.039	2.0	ug/L	192	Standard
	Cu	65	754.7	3.1	**0.3255**	0.008	2.4	ug/L	141	Standard
	Zn	66	4017.9	2.4	**3.0152**	0.057	1.9	ug/L	491	Standard
>	**Ge**	72	491414.0	2.7				ug/L	515087	Standard
	As	75	1051.2	2.3	**1.0210**	0.045	4.4	ug/L	20	Standard
	Se	82	143.6	4.5	**1.1879**	0.030	2.5	ug/L	15	Standard
	Se-1	77	105.7	15.2	**0.7547**	0.249	32.9	ug/L	48	Standard
>	**Ga**	71	188.3	17.7				mg/L	90	Standard
	Rb	85	950.0	0.9				ug/L	23	Standard
	Y	89	424439.8	0.9				ug/L	448510	Standard
>	**Rh**	103	96.7	7.9				ug/L	15	Standard
	Mo	98	5849.5	0.5	**2.2126**	0.029	1.3	ug/L	11	Standard
	Ag	107	89.7	21.1	**-0.0024**	0.003	128.8	ug/L	134	Standard
	Cd	111	11.8	58.9	**0.0028**	0.003	121.4	mg/L	7	Standard
	Cd	114	46.1	44.3	**0.0042**	0.004	96.2	ug/L	16	Standard
>	**In**	115	592706.9	1.7				ug/L	611117	Standard
	Sn	118	891.7	9.5	**0.0535**	0.013	23.6	ug/L	518	Standard
	Sb	123	236.5	14.9	**0.0388**	0.007	18.8	ug/L	127	Standard
	Ba	135	171509.6	0.5	**85.6510**	1.837	2.1	ug/L	45	Standard
	Ce	140	703.3	3.9				ug/L	298	Standard
>	**Tb**	159	954495.6	1.0				ug/L	977942	Standard
	Ho	165	26.7	60.3				ug/L	12	Standard
	Tl	203	128.7	6.5	**0.0108**	0.001	9.4	ug/L	31	Standard
	Tl	205	116.7	2.5	**0.0140**	0.000	2.9	ug/L	20	Standard
	Pb	206	518.0	5.9	**0.0341**	0.006	17.0	ug/L	286	Standard
	Pb	207	434.7	5.8	**0.0328**	0.005	16.1	ug/L	257	Standard
	Pb	208	1843.7	3.4	**0.0392**	0.003	8.9	ug/L	956	Standard
	U	238	11694.2	0.7	**1.6602**	0.012	0.7	ug/L	17	Standard
>	**Bi**	209	546693.4	0.1				ug/L	557080	Standard

Sample ID: L1606115501
Report Date/Time: Tuesday, June 28, 2016 17:18:31
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.9680**	18.988	173.1	mg/L	0	Standard
Mg	24	30276.9	1.0	**40.7793**	0.477	1.2	mg/L	32	Standard
K	39	240.0	12.5	**0.6798**	0.099	14.5	mg/L	30	Standard
Ca	43	255.0	15.3	**33.9356**	6.415	18.9	mg/L	48	Standard
Fe	54	184.3	20.5	**0.0318**	0.034	107.2	mg/L	159	Standard
Fe	57	341.7	3.0	**0.5441**	0.035	6.5	mg/L	195	Standard
> **Sc-1**	45	39028.5	1.1				mg/L	39543	Standard
Cl	35	37440.9	1.2				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	32190.8	1.7				ug/L	1647	Standard
P	31	22541.8	0.7				ug/L	15231	Standard
S	34	2386.9	3.2				ug/L	1845	Standard
Sr	88	303.3	20.6				ug/L	70	Standard
C	12	480.0	6.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	27.6	75.9				mg/L	12	Standard
Ho-1	165	26.7	60.3				mg/L	12	Standard
Er	166	50.0	40.0				mg/L	20	Standard
I	127	1264643.9	6.0				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		113.247	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		95.404	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606115501
Report Date/Time: Tuesday, June 28, 2016 17:18:31
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.988
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.135
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115502
Sample Date/Time: Tuesday, June 28, 2016 17:19:26
Number of Replicates: 3
Autosampler Position: 253
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94157.1	4.7				ug/L	81760	Standard
	Be	9	20.0	50.0	**0.0078**	0.009	115.3	ug/L	17	Standard
	Al	27	28931934.4	4.0	**163.0048**	1.392	0.9	ug/L	1410	Standard
	Sc	45	40320.3	3.5				ug/L	39543	Standard
	Ti	47	168.3	17.1	**0.5232**	0.117	22.4	ug/L	39	Standard
	V	51	1551.4	10.5	**0.0205**	0.027	130.5	ug/L	1415	Standard
	Cr	52	10328.9	2.0	**0.4515**	0.024	5.3	ug/L	6946	Standard
	Cr	53	2665.2	7.9	**2.7435**	0.199	7.2	ug/L	328	Standard
	Mn	55	1534029.1	1.0	**240.4438**	3.129	1.3	ug/L	823	Standard
	Co	59	1102.0	2.8	**0.1281**	0.004	3.4	ug/L	252	Standard
	Ni	60	1786.8	2.8	**0.8597**	0.029	3.3	ug/L	192	Standard
	Cu	65	955.7	1.4	**0.4206**	0.012	3.0	ug/L	141	Standard
	Zn	66	3611.4	2.8	**2.5229**	0.115	4.5	ug/L	491	Standard
>	**Ge**	72	506659.2	1.8				ug/L	515087	Standard
	As	75	1883.3	1.6	**1.7433**	0.024	1.4	ug/L	20	Standard
	Se	82	155.4	0.8	**1.2536**	0.030	2.4	ug/L	15	Standard
	Se-1	77	168.7	6.6	**1.5590**	0.181	11.6	ug/L	48	Standard
>	**Ga**	71	140.0	15.6				mg/L	90	Standard
	Rb	85	1435.1	8.2				ug/L	23	Standard
	Y	89	449200.8	1.5				ug/L	448510	Standard
>	**Rh**	103	93.3	16.4				ug/L	15	Standard
	Mo	98	2699.1	0.6	**0.9667**	0.027	2.8	ug/L	11	Standard
	Ag	107	94.7	12.7	**-0.0023**	0.002	96.2	ug/L	134	Standard
	Cd	111	15.7	16.0	**0.0044**	0.001	31.4	mg/L	7	Standard
	Cd	114	28.9	55.3	**0.0006**	0.003	513.8	ug/L	16	Standard
>	**In**	115	619397.8	3.1				ug/L	611117	Standard
	Sn	118	1035.0	8.9	**0.0707**	0.012	17.4	ug/L	518	Standard
	Sb	123	177.5	18.8	**0.0244**	0.008	32.9	ug/L	127	Standard
	Ba	135	342505.9	1.5	**163.7121**	2.675	1.6	ug/L	45	Standard
	Ce	140	470.0	1.8				ug/L	298	Standard
>	**Tb**	159	978535.3	2.3				ug/L	977942	Standard
	Ho	165	33.3	17.3				ug/L	12	Standard
	Tl	203	95.7	2.4	**0.0067**	0.000	2.7	ug/L	31	Standard
	Tl	205	98.3	7.8	**0.0112**	0.001	8.1	ug/L	20	Standard
	Pb	206	616.7	3.3	**0.0513**	0.003	5.5	ug/L	286	Standard
	Pb	207	520.0	3.8	**0.0494**	0.003	5.9	ug/L	257	Standard
	Pb	208	2224.4	1.1	**0.0583**	0.001	1.4	ug/L	956	Standard
	U	238	39.7	29.5	**0.0026**	0.002	61.0	ug/L	17	Standard
>	**Bi**	209	553220.4	1.1				ug/L	557080	Standard

Sample ID: L1606115502
Report Date/Time: Tuesday, June 28, 2016 17:21:43
Page 1

Approved: June 29, 2016

Na	23	5.0	0.0	**31.5109**	1.087	3.4	mg/L	0	Standard
Mg	24	25140.9	3.4	**32.7721**	0.865	2.6	mg/L	32	Standard
K	39	73.3	23.9	**0.1023**	0.063	61.8	mg/L	30	Standard
Ca	43	243.3	6.6	**30.9203**	2.535	8.2	mg/L	48	Standard
Fe	54	2629.1	2.7	**2.1972**	0.059	2.7	mg/L	159	Standard
Fe	57	1076.7	7.2	**2.9489**	0.380	12.9	mg/L	195	Standard
> **Sc-1**	45	40320.3	3.5				mg/L	39543	Standard
Cl	35	39998.8	2.6				ug/L	27642	Standard
Kr	83	2.7	43.3				ug/L	2	Standard
Br	81	32989.2	3.3				ug/L	1647	Standard
P	31	24089.1	1.8				ug/L	15231	Standard
S	34	2583.6	5.0				ug/L	1845	Standard
Sr	88	433.3	4.7				ug/L	70	Standard
C	12	350.0	5.7				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	25.7	20.9				mg/L	12	Standard
Ho-1	165	33.3	17.3				mg/L	12	Standard
Er	166	20.0	50.0				mg/L	20	Standard
I	127	377170.1	0.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		115.163	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.364	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606115502
Report Date/Time: Tuesday, June 28, 2016 17:21:43
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	101.355
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	99.307
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606115502
Report Date/Time: Tuesday, June 28, 2016 17:21:43
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606115503
Sample Date/Time: Tuesday, June 28, 2016 17:22:37
Number of Replicates: 3
Autosampler Position: 254
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93460.7	0.5				ug/L	81760	Standard
	Be	9	35.0	14.3	**0.0211**	0.004	21.2	ug/L	17	Standard
	Al	27	14187081.3	1.7	**80.5012**	1.275	1.6	ug/L	1410	Standard
	Sc	45	39556.6	4.7				ug/L	39543	Standard
	Ti	47	229.0	18.7	**0.7435**	0.168	22.5	ug/L	39	Standard
	V	51	2172.0	8.1	**0.1016**	0.022	21.3	ug/L	1415	Standard
	Cr	52	10167.4	1.5	**0.3846**	0.034	8.8	ug/L	6946	Standard
	Cr	53	1090.0	15.2	**0.8840**	0.185	20.9	ug/L	328	Standard
	Mn	55	1432250.3	1.3	**217.9180**	5.450	2.5	ug/L	823	Standard
	Co	59	795.7	6.2	**0.0804**	0.008	9.7	ug/L	252	Standard
	Ni	60	1574.1	4.4	**0.7198**	0.033	4.6	ug/L	192	Standard
	Cu	65	1999.1	3.4	**0.9475**	0.036	3.8	ug/L	141	Standard
	Zn	66	5797.8	1.6	**4.3927**	0.080	1.8	ug/L	491	Standard
>	**Ge**	72	521945.3	1.2				ug/L	515087	Standard
	As	75	168.7	6.2	**0.1883**	0.009	4.6	ug/L	20	Standard
	Se	82	81.0	7.2	**0.5749**	0.055	9.6	ug/L	15	Standard
	Se-1	77	72.3	14.9	**0.2276**	0.131	57.6	ug/L	48	Standard
>	**Ga**	71	155.0	23.3				mg/L	90	Standard
	Rb	85	10031.7	2.6				ug/L	23	Standard
	Y	89	435099.7	2.5				ug/L	448510	Standard
>	**Rh**	103	70.0	28.6				ug/L	15	Standard
	Mo	98	145.2	5.2	**0.0354**	0.003	8.7	ug/L	11	Standard
	Ag	107	85.3	8.5	**-0.0033**	0.001	30.7	ug/L	134	Standard
	Cd	111	76.7	2.3	**0.0335**	0.001	2.5	mg/L	7	Standard
	Cd	114	203.1	19.3	**0.0347**	0.008	22.4	ug/L	16	Standard
>	**In**	115	602973.5	0.6				ug/L	611117	Standard
	Sn	118	941.7	4.9	**0.0595**	0.009	14.6	ug/L	518	Standard
	Sb	123	323.8	6.4	**0.0567**	0.004	7.5	ug/L	127	Standard
	Ba	135	119338.2	0.9	**58.5639**	0.899	1.5	ug/L	45	Standard
	Ce	140	855.0	10.9				ug/L	298	Standard
>	**Tb**	159	952983.8	1.5				ug/L	977942	Standard
	Ho	165	36.7	41.7				ug/L	12	Standard
	Tl	203	88.7	6.9	**0.0057**	0.001	14.0	ug/L	31	Standard
	Tl	205	55.0	27.3	**0.0052**	0.002	38.2	ug/L	20	Standard
	Pb	206	861.0	5.9	**0.0956**	0.010	10.1	ug/L	286	Standard
	Pb	207	708.7	5.0	**0.0873**	0.008	9.7	ug/L	257	Standard
	Pb	208	2964.1	2.3	**0.0967**	0.005	5.3	ug/L	956	Standard
	U	238	696.7	2.0	**0.0942**	0.003	2.9	ug/L	17	Standard
>	**Bi**	209	557355.1	1.0				ug/L	557080	Standard

Sample ID: L1606115503
Report Date/Time: Tuesday, June 28, 2016 17:24:54
Page 1

Approved: June 29, 2016

Na	23	6.7	43.3	**42.2849**	16.098	38.1	mg/L	0	Standard
Mg	24	19544.3	4.8	**25.9961**	1.909	7.3	mg/L	32	Standard
K	39	113.3	15.5	**0.2405**	0.047	19.5	mg/L	30	Standard
Ca	43	286.7	25.9	**38.5627**	12.897	33.4	mg/L	48	Standard
Fe	54	997.4	5.8	**0.7664**	0.064	8.4	mg/L	159	Standard
Fe	57	530.0	13.9	**1.1674**	0.275	23.5	mg/L	195	Standard
Sc-1	45	39556.6	4.7				mg/L	39543	Standard
Cl	35	38409.3	3.3				ug/L	27642	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	17249.8	6.1				ug/L	1647	Standard
P	31	22414.9	3.0				ug/L	15231	Standard
S	34	2625.2	3.4				ug/L	1845	Standard
Sr	88	228.3	6.7				ug/L	70	Standard
C	12	260.0	15.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	21.4	67.1				mg/L	12	Standard
Ho-1	165	36.7	41.7				mg/L	12	Standard
Er	166	40.0	75.0				mg/L	20	Standard
I	127	663837.8	8.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		114.311	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		101.332	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606115503
Report Date/Time: Tuesday, June 28, 2016 17:24:54
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.668
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.049
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 17:25:50
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	91554.4	4.0				ug/L	81760	Standard
	Be	9	59302.5	0.5	**53.3655**	1.843	3.5	ug/L	17	Standard
	Al	27	8323459.6	1.6	**48.2773**	2.634	5.5	ug/L	1410	Standard
	Sc	45	40460.6	1.5				ug/L	39543	Standard
	Ti	47	25609.3	2.1	**100.1533**	1.793	1.8	ug/L	39	Standard
	V	51	364470.3	1.4	**50.8781**	0.853	1.7	ug/L	1415	Standard
	Cr	52	366105.1	1.7	**50.1771**	0.675	1.3	ug/L	6946	Standard
	Cr	53	45260.8	0.4	**49.8893**	0.509	1.0	ug/L	328	Standard
	Mn	55	332731.4	2.3	**49.8507**	1.205	2.4	ug/L	823	Standard
	Co	59	382266.4	2.9	**52.9683**	1.437	2.7	ug/L	252	Standard
	Ni	60	97387.3	1.2	**50.2694**	0.545	1.1	ug/L	192	Standard
	Cu	65	98831.3	1.3	**50.5786**	0.557	1.1	ug/L	141	Standard
	Zn	66	57834.9	1.6	**50.5762**	0.877	1.7	ug/L	491	Standard
>	Ge	72	528768.9	1.2				ug/L	515087	Standard
	As	75	58724.5	1.2	**50.9203**	0.224	0.4	ug/L	20	Standard
	Se	82	6151.0	1.0	**51.9261**	0.245	0.5	ug/L	15	Standard
	Se-1	77	4089.2	2.2	**52.1522**	0.563	1.1	ug/L	48	Standard
>	Ga	71	85.0	21.2				mg/L	90	Standard
	Rb	85	851.7	6.5				ug/L	23	Standard
	Y	89	456324.8	0.4				ug/L	448510	Standard
>	Rh	103	36.7	39.4				ug/L	15	Standard
	Mo	98	284202.4	1.9	**102.0947**	3.648	3.6	ug/L	11	Standard
	Ag	107	351396.4	1.1	**50.8197**	1.435	2.8	ug/L	134	Standard
	Cd	111	110732.8	1.5	**50.3025**	1.470	2.9	mg/L	7	Standard
	Cd	114	271688.9	2.3	**50.7793**	1.915	3.8	ug/L	16	Standard
>	In	115	629514.4	1.7				ug/L	611117	Standard
	Sn	118	310739.2	0.7	**50.9245**	0.924	1.8	ug/L	518	Standard
	Sb	123	246166.6	1.6	**50.6553**	1.659	3.3	ug/L	127	Standard
	Ba	135	104825.0	0.9	**49.2805**	1.265	2.6	ug/L	45	Standard
	Ce	140	70.0	55.8				ug/L	298	Standard
>	Tb	159	987475.2	1.7				ug/L	977942	Standard
	Ho	165	5.0	100.0				ug/L	12	Standard
	Tl	203	444397.7	1.4	**50.9180**	1.232	2.4	ug/L	31	Standard
	Tl	205	396866.3	1.6	**52.3722**	1.695	3.2	ug/L	20	Standard
	Pb	206	284750.7	1.1	**50.8069**	1.281	2.5	ug/L	286	Standard
	Pb	207	259280.5	1.3	**51.2940**	1.414	2.8	ug/L	257	Standard
	Pb	208	1040337.0	1.7	**53.1998**	1.656	3.1	ug/L	956	Standard
	U	238	390662.1	1.3	**52.6405**	1.349	2.6	ug/L	17	Standard
>	Bi	209	577126.2	1.6				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:28:07
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4445.7	6.0	**5.7289**	0.397	6.9	mg/L	32	Standard
K	39	1320.1	5.5	**4.2194**	0.276	6.5	mg/L	30	Standard
Ca	43	81.7	24.7	**6.7579**	2.904	43.0	mg/L	48	Standard
Fe	54	5927.1	5.8	**5.1036**	0.243	4.8	mg/L	159	Standard
Fe	57	1661.8	3.7	**4.8569**	0.121	2.5	mg/L	195	Standard
Sc-1	45	40460.6	1.5				mg/L	39543	Standard
Cl	35	38764.2	1.7				ug/L	27642	Standard
Kr	83	2.3	107.9				ug/L	2	Standard
Br	81	2123.5	9.4				ug/L	1647	Standard
P	31	21915.8	3.2				ug/L	15231	Standard
S	34	2866.9	6.9				ug/L	1845	Standard
Sr	88	56.7	10.2				ug/L	70	Standard
C	12	170.0	44.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	173.2				mg/L	0	Standard
Dy	164	9.7	103.3				mg/L	12	Standard
Ho-1	165	5.0	100.0				mg/L	12	Standard
Er	166	6.7	86.6				mg/L	20	Standard
I	127	21360.9	50.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	106.731		
Al	27	96.555		
Sc	45			
Ti	47	100.153		
V	51	101.756		
Cr	52	100.354		
Cr	53			
Mn	55	99.701		
Co	59	105.937		
Ni	60	100.539		
Cu	65	101.157		
Zn	66	101.152		
Ge	72		102.656	
As	75	101.841		
Se	82	103.852		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:28:07
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	102.095		
Ag	107	101.639		
Cd	111	100.605		
Cd	114			
> In	115			103.011
Sn	118	101.849		
Sb	123	101.311		
Ba	135	98.561		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.836		
Tl	205			
Pb	206			
Pb	207			
Pb	208	106.400		
U	238	105.281		
> Bi	209			103.598
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 17:28:07
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 17:29:02
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89289.0	1.1				ug/L	81760	Standard
	Be	9	35.0	51.5	**0.0225**	0.017	74.1	ug/L	17	Standard
	Al	27	14272.5	142.7	**0.0706**	0.122	172.0	ug/L	1410	Standard
	Sc	45	39427.9	1.8				ug/L	39543	Standard
	Ti	47	28.0	28.3	**-0.0534**	0.032	59.3	ug/L	39	Standard
	V	51	1205.0	11.4	**-0.0346**	0.020	56.4	ug/L	1415	Standard
	Cr	52	6173.6	1.1	**-0.1773**	0.009	5.0	ug/L	6946	Standard
	Cr	53	500.0	6.0	**0.2236**	0.034	15.3	ug/L	328	Standard
	Mn	55	1435.1	83.0	**0.0711**	0.182	256.2	ug/L	823	Standard
	Co	59	252.7	26.0	**0.0043**	0.009	214.7	ug/L	252	Standard
	Ni	60	265.7	8.4	**0.0350**	0.012	33.5	ug/L	192	Standard
	Cu	65	144.0	13.3	**-0.0157**	0.010	63.6	ug/L	141	Standard
	Zn	66	428.7	4.3	**-0.4396**	0.017	3.8	ug/L	491	Standard
>	**Ge**	72	519199.6	0.1				ug/L	515087	Standard
	As	75	-13.2	109.3	**0.0287**	0.013	44.2	ug/L	20	Standard
	Se	82	24.7	35.1	**0.0933**	0.075	79.8	ug/L	15	Standard
	Se-1	77	53.0	16.4	**-0.0209**	0.115	551.4	ug/L	48	Standard
>	**Ga**	71	56.7	41.7				mg/L	90	Standard
	Rb	85	13.3	78.1				ug/L	23	Standard
	Y	89	435848.4	0.6				ug/L	448510	Standard
>	**Rh**	103	16.7	45.8				ug/L	15	Standard
	Mo	98	252.5	21.9	**0.0755**	0.022	29.1	ug/L	11	Standard
	Ag	107	123.0	3.3	**0.0023**	0.000	14.7	ug/L	134	Standard
	Cd	111	11.8	114.0	**0.0027**	0.006	236.7	mg/L	7	Standard
	Cd	114	47.6	83.9	**0.0044**	0.008	178.8	ug/L	16	Standard
>	**In**	115	604937.7	1.9				ug/L	611117	Standard
	Sn	118	835.0	26.8	**0.0403**	0.035	87.6	ug/L	518	Standard
	Sb	123	438.3	28.1	**0.0807**	0.025	30.6	ug/L	127	Standard
	Ba	135	40.7	30.0	**-0.0052**	0.006	117.6	ug/L	45	Standard
	Ce	140	35.0	37.8				ug/L	298	Standard
>	**Tb**	159	968167.7	1.8				ug/L	977942	Standard
	Ho	165	8.3	124.9				ug/L	12	Standard
	Tl	203	65.7	40.9	**0.0028**	0.003	111.1	ug/L	31	Standard
	Tl	205	198.3	119.2	**0.0243**	0.032	131.1	ug/L	20	Standard
	Pb	206	312.0	4.9	**-0.0071**	0.003	48.0	ug/L	286	Standard
	Pb	207	265.7	2.9	**-0.0047**	0.002	47.8	ug/L	257	Standard
	Pb	208	1274.7	38.3	**0.0057**	0.026	450.9	ug/L	956	Standard
	U	238	172.3	130.6	**0.0207**	0.031	149.9	ug/L	17	Standard
>	**Bi**	209	569803.2	1.6				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:31:19
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**21.6890**	18.779	86.6	mg/L	0	Standard
Mg	24	61.7	47.5	**0.0256**	0.040	158.0	mg/L	32	Standard
K	39	26.7	10.8	**-0.0510**	0.010	20.5	mg/L	30	Standard
Ca	43	21.7	48.0	**-2.0553**	1.552	75.5	mg/L	48	Standard
Fe	54	129.4	21.4	**-0.0198**	0.024	119.8	mg/L	159	Standard
Fe	57	196.7	34.5	**0.0395**	0.221	559.0	mg/L	195	Standard
Sc-1	45	39427.9	1.8				mg/L	39543	Standard
Cl	35	37196.3	1.0				ug/L	27642	Standard
Kr	83	2.7	21.7				ug/L	2	Standard
Br	81	1853.4	7.9				ug/L	1647	Standard
P	31	20680.8	2.7				ug/L	15231	Standard
S	34	2700.2	3.5				ug/L	1845	Standard
Sr	88	68.3	23.5				ug/L	70	Standard
C	12	133.3	38.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.5	107.6				mg/L	12	Standard
Ho-1	165	8.3	124.9				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	7115.0	15.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		100.798	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:31:19
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.989
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.284
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 17:31:19
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606114201
Sample Date/Time: Tuesday, June 28, 2016 18:46:32
Number of Replicates: 3
Autosampler Position: 255
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78164.5	2.4				ug/L	81760	Standard
	Be	9	18.3	41.7	**0.0097**	0.008	86.8	ug/L	17	Standard
	Al	27	121659.9	1.7	**0.8116**	0.028	3.5	ug/L	1410	Standard
	Sc	45	40555.9	2.8				ug/L	39543	Standard
	Ti	47	55.7	15.0	**0.0583**	0.036	61.2	ug/L	39	Standard
	V	51	1460.1	6.7	**0.0028**	0.012	438.4	ug/L	1415	Standard
	Cr	52	7975.8	0.8	**0.0855**	0.018	21.3	ug/L	6946	Standard
	Cr	53	690.0	12.1	**0.4421**	0.091	20.5	ug/L	328	Standard
	Mn	55	73452.6	0.4	**11.1499**	0.088	0.8	ug/L	823	Standard
	Co	59	2365.9	3.2	**0.3044**	0.010	3.2	ug/L	252	Standard
	Ni	60	956.0	2.8	**0.4012**	0.010	2.4	ug/L	192	Standard
	Cu	65	203.7	5.3	**0.0160**	0.005	33.1	ug/L	141	Standard
	Zn	66	1994.1	0.1	**0.9866**	0.018	1.8	ug/L	491	Standard
>	**Ge**	72	516559.4	1.0				ug/L	515087	Standard
	As	75	-28.9	32.9	**0.0147**	0.008	56.4	ug/L	20	Standard
	Se	82	19.3	15.0	**0.0479**	0.026	53.5	ug/L	15	Standard
	Se-1	77	59.0	13.2	**0.0614**	0.097	157.2	ug/L	48	Standard
>	**Ga**	71	36.7	51.6				mg/L	90	Standard
	Rb	85	463.3	2.5				ug/L	23	Standard
	Y	89	448462.9	2.6				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	65.9	7.4	**0.0053**	0.002	42.6	ug/L	11	Standard
	Ag	107	82.7	4.2	**-0.0039**	0.001	16.6	ug/L	134	Standard
	Cd	111	8.9	29.8	**0.0012**	0.001	94.7	mg/L	7	Standard
	Cd	114	20.4	73.3	**-0.0010**	0.003	278.6	ug/L	16	Standard
>	**In**	115	611220.6	1.9				ug/L	611117	Standard
	Sn	118	670.0	14.0	**0.0115**	0.017	148.1	ug/L	518	Standard
	Sb	123	214.2	30.6	**0.0326**	0.014	43.4	ug/L	127	Standard
	Ba	135	7677.9	3.5	**3.6947**	0.165	4.5	ug/L	45	Standard
	Ce	140	396.7	10.7				ug/L	298	Standard
>	**Tb**	159	963569.5	0.7				ug/L	977942	Standard
	Ho	165	16.7	45.8				ug/L	12	Standard
	Tl	203	21.7	17.5	**-0.0022**	0.000	19.5	ug/L	31	Standard
	Tl	205	13.3	21.7	**-0.0005**	0.000	79.8	ug/L	20	Standard
	Pb	206	368.7	4.6	**0.0040**	0.003	86.3	ug/L	286	Standard
	Pb	207	283.0	4.1	**-0.0005**	0.002	459.4	ug/L	257	Standard
	Pb	208	1139.7	3.0	**-0.0006**	0.002	246.5	ug/L	956	Standard
	U	238	10.3	39.1	**-0.0015**	0.001	36.3	ug/L	17	Standard
>	**Bi**	209	562456.7	0.5				ug/L	557080	Standard

Sample ID: L1606114201
Report Date/Time: Tuesday, June 28, 2016 18:48:49
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	763.4	13.7	**0.9351**	0.146	15.6	mg/L	32	Standard
K	39	31.7	39.7	**-0.0363**	0.045	123.6	mg/L	30	Standard
Ca	43	18.3	31.5	**-2.6404**	0.796	30.1	mg/L	48	Standard
Fe	54	359.1	17.1	**0.1792**	0.047	26.3	mg/L	159	Standard
Fe	57	273.3	13.7	**0.2778**	0.149	53.6	mg/L	195	Standard
> **Sc-1**	45	40555.9	2.8				mg/L	39543	Standard
Cl	35	37427.5	0.7				ug/L	27642	Standard
Kr	83	0.0					ug/L	2	Standard
Br	81	2536.9	6.5				ug/L	1647	Standard
P	31	19003.6	0.9				ug/L	15231	Standard
S	34	2375.2	4.8				ug/L	1845	Standard
Sr	88	91.7	13.7				ug/L	70	Standard
C	12	86.7	29.0				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	29.2	34.3				mg/L	12	Standard
Ho-1	165	16.7	45.8				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	22815.7	12.2				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		95.602	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		100.286	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606114201
Report Date/Time: Tuesday, June 28, 2016 18:48:49
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.017
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.965
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 18:49:46
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	77476.0	3.1				ug/L	81760	Standard
	Be	9	45015.2	3.8	**47.8878**	3.193	6.7	ug/L	17	Standard
	Al	27	6953697.9	2.2	**47.6359**	2.288	4.8	ug/L	1410	Standard
	Sc	45	38896.5	2.1				ug/L	39543	Standard
	Ti	47	23530.3	2.9	**93.4171**	3.639	3.9	ug/L	39	Standard
	V	51	333406.5	0.5	**47.2267**	0.631	1.3	ug/L	1415	Standard
	Cr	52	337441.2	2.6	**46.8861**	1.719	3.7	ug/L	6946	Standard
	Cr	53	42128.5	3.1	**47.1231**	1.874	4.0	ug/L	328	Standard
	Mn	55	311025.1	1.6	**47.2965**	1.233	2.6	ug/L	823	Standard
	Co	59	333541.6	0.3	**46.9106**	0.523	1.1	ug/L	252	Standard
	Ni	60	90766.5	2.7	**47.5598**	1.737	3.7	ug/L	192	Standard
	Cu	65	91363.1	2.5	**47.4629**	1.646	3.5	ug/L	141	Standard
	Zn	66	53283.5	1.9	**47.2487**	1.367	2.9	ug/L	491	Standard
>	Ge	72	520941.4	1.0				ug/L	515087	Standard
	As	75	53453.8	0.5	**47.0530**	0.571	1.2	ug/L	20	Standard
	Se	82	5409.0	0.9	**46.3406**	0.883	1.9	ug/L	15	Standard
	Se-1	77	3695.1	1.5	**47.7773**	0.360	0.8	ug/L	48	Standard
>	Ga	71	58.3	27.6				mg/L	90	Standard
	Rb	85	731.7	6.2				ug/L	23	Standard
	Y	89	433135.5	0.5				ug/L	448510	Standard
>	Rh	103	21.7	13.3				ug/L	15	Standard
	Mo	98	251705.5	0.6	**94.1208**	1.331	1.4	ug/L	11	Standard
	Ag	107	314845.2	1.8	**47.3955**	0.605	1.3	ug/L	134	Standard
	Cd	111	98007.4	2.2	**46.3425**	0.750	1.6	mg/L	7	Standard
	Cd	114	239351.7	3.2	**46.5659**	1.544	3.3	ug/L	16	Standard
>	In	115	604583.1	1.4				ug/L	611117	Standard
	Sn	118	276455.2	2.3	**47.1543**	0.498	1.1	ug/L	518	Standard
	Sb	123	224560.9	2.5	**48.0923**	0.641	1.3	ug/L	127	Standard
	Ba	135	94923.9	3.0	**46.4488**	1.094	2.4	ug/L	45	Standard
	Ce	140	80.0	22.5				ug/L	298	Standard
>	Tb	159	958457.8	2.2				ug/L	977942	Standard
	Ho	165	13.3	21.7				ug/L	12	Standard
	Tl	203	399564.4	2.5	**47.4112**	1.332	2.8	ug/L	31	Standard
	Tl	205	347775.3	2.0	**47.5216**	1.083	2.3	ug/L	20	Standard
	Pb	206	259967.8	1.3	**48.0301**	0.722	1.5	ug/L	286	Standard
	Pb	207	236539.5	0.9	**48.4530**	0.488	1.0	ug/L	257	Standard
	Pb	208	913309.7	2.1	**48.3569**	1.176	2.4	ug/L	956	Standard
	U	238	337839.3	1.6	**47.1413**	0.910	1.9	ug/L	17	Standard
>	Bi	209	557186.5	0.3				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 18:52:03
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3633.8	7.1	**4.8666**	0.455	9.3	mg/L	32	Standard
K	39	1313.4	12.2	**4.3752**	0.599	13.7	mg/L	30	Standard
Ca	43	51.7	20.1	**2.6371**	1.625	61.6	mg/L	48	Standard
Fe	54	5289.1	7.8	**4.7357**	0.483	10.2	mg/L	159	Standard
Fe	57	1531.7	10.1	**4.6389**	0.613	13.2	mg/L	195	Standard
> **Sc-1**	45	38896.5	2.1				mg/L	39543	Standard
Cl	35	35022.4	1.4				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	1490.1	8.7				ug/L	1647	Standard
P	31	19075.3	3.9				ug/L	15231	Standard
S	34	2581.9	1.6				ug/L	1845	Standard
Sr	88	70.0	18.9				ug/L	70	Standard
C	12	143.3	33.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	6.0	193.9				mg/L	12	Standard
Ho-1	165	13.3	21.7				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	2193.5	28.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	95.776		
Al	27	95.272		
Sc	45			
Ti	47	93.417		
V	51	94.453		
Cr	52	93.772		
Cr	53			
Mn	55	94.593		
Co	59	93.821		
Ni	60	95.120		
Cu	65	94.926		
Zn	66	94.497		
> Ge	72		101.137	
As	75	94.106		
Se	82	92.681		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 18:52:03
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	94.121		
Ag	107	94.791		
Cd	111	92.685		
Cd	114			
> In	115			98.931
Sn	118	94.309		
Sb	123	96.185		
Ba	135	92.898		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	94.822		
Tl	205			
Pb	206			
Pb	207			
Pb	208	96.714		
U	238	94.283		
> Bi	209			100.019
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 18:52:03
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 18:52:57
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75264.9	3.1				ug/L	81760	Standard
	Be	9	28.3	44.4	**0.0211**	0.013	62.2	ug/L	17	Standard
	Al	27	1848.4	7.3	**-0.0014**	0.001	57.2	ug/L	1410	Standard
	Sc	45	39733.8	3.9				ug/L	39543	Standard
	Ti	47	30.7	21.2	**-0.0408**	0.026	64.0	ug/L	39	Standard
	V	51	1157.9	3.5	**-0.0386**	0.007	18.1	ug/L	1415	Standard
	Cr	52	6209.6	2.0	**-0.1578**	0.015	9.7	ug/L	6946	Standard
	Cr	53	373.3	6.7	**0.0873**	0.028	32.1	ug/L	328	Standard
	Mn	55	647.3	1.5	**-0.0478**	0.002	4.7	ug/L	823	Standard
	Co	59	232.0	9.1	**0.0019**	0.003	163.0	ug/L	252	Standard
	Ni	60	166.3	9.0	**-0.0160**	0.007	44.1	ug/L	192	Standard
	Cu	65	124.7	14.0	**-0.0248**	0.008	34.1	ug/L	141	Standard
	Zn	66	405.0	8.1	**-0.4552**	0.029	6.3	ug/L	491	Standard
>	**Ge**	72	510975.0	1.4				ug/L	515087	Standard
	As	75	-43.1	41.4	**0.0017**	0.016	900.9	ug/L	20	Standard
	Se	82	19.6	25.0	**0.0525**	0.045	86.4	ug/L	15	Standard
	Se-1	77	59.7	5.9	**0.0797**	0.050	62.7	ug/L	48	Standard
>	**Ga**	71	36.7	39.4				mg/L	90	Standard
	Rb	85	23.3	32.7				ug/L	23	Standard
	Y	89	435979.5	2.5				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	250.3	25.6	**0.0730**	0.024	32.3	ug/L	11	Standard
	Ag	107	100.0	11.8	**-0.0014**	0.002	125.8	ug/L	134	Standard
	Cd	111	11.9	34.0	**0.0026**	0.002	71.6	mg/L	7	Standard
	Cd	114	45.3	6.7	**0.0038**	0.000	13.0	ug/L	16	Standard
>	**In**	115	614932.8	1.3				ug/L	611117	Standard
	Sn	118	795.0	23.6	**0.0316**	0.031	97.7	ug/L	518	Standard
	Sb	123	431.0	25.7	**0.0779**	0.023	29.6	ug/L	127	Standard
	Ba	135	30.3	24.3	**-0.0105**	0.003	32.1	ug/L	45	Standard
	Ce	140	20.0	66.1				ug/L	298	Standard
>	**Tb**	159	946813.4	1.4				ug/L	977942	Standard
	Ho	165	11.7	89.2				ug/L	12	Standard
	Tl	203	48.3	29.0	**0.0009**	0.002	172.8	ug/L	31	Standard
	Tl	205	35.0	86.9	**0.0025**	0.004	165.9	ug/L	20	Standard
	Pb	206	284.0	5.8	**-0.0111**	0.003	26.2	ug/L	286	Standard
	Pb	207	257.7	11.7	**-0.0053**	0.006	110.9	ug/L	257	Standard
	Pb	208	990.3	3.8	**-0.0080**	0.002	20.6	ug/L	956	Standard
	U	238	43.0	24.2	**0.0030**	0.001	48.0	ug/L	17	Standard
>	**Bi**	209	557905.1	0.6				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 18:55:14
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	31.7	18.2	**-0.0149**	0.009	60.8	mg/L	32	Standard
K	39	20.0	50.0	**-0.0742**	0.034	46.2	mg/L	30	Standard
Ca	43	35.0	62.3	**-0.0114**	3.447	30330.8	mg/L	48	Standard
Fe	54	144.1	21.6	**-0.0065**	0.032	498.8	mg/L	159	Standard
Fe	57	238.3	3.2	**0.1775**	0.054	30.7	mg/L	195	Standard
Sc-1	45	39733.8	3.9				mg/L	39543	Standard
Cl	35	35783.5	1.7				ug/L	27642	Standard
Kr	83	3.3	45.8				ug/L	2	Standard
Br	81	1466.7	6.9				ug/L	1647	Standard
P	31	19612.7	1.4				ug/L	15231	Standard
S	34	2756.9	2.6				ug/L	1845	Standard
Sr	88	66.7	37.0				ug/L	70	Standard
C	12	86.7	6.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	32.7	47.4				mg/L	12	Standard
Ho-1	165	11.7	89.2				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	1593.4	21.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.202	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 18:55:14
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.624
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.148
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 18:55:14
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Tuesday, June 28, 2016 18:56:10
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	71096.5	2.9				ug/L	81760	Standard
	Be	9	180.0	15.5	**0.1982**	0.026	13.3	ug/L	17	Standard
	Al	27	1130.0	11.5	**-0.0060**	0.001	12.9	ug/L	1410	Standard
	Sc	45	37422.8	0.9				ug/L	39543	Standard
	Ti	47	25.0	10.6	**-0.0625**	0.010	16.0	ug/L	39	Standard
	V	51	3508.6	1.4	**0.3092**	0.005	1.6	ug/L	1415	Standard
	Cr	52	11162.1	1.4	**0.5814**	0.031	5.3	ug/L	6946	Standard
	Cr	53	975.0	4.9	**0.7938**	0.066	8.3	ug/L	328	Standard
	Mn	55	3454.1	1.9	**0.3960**	0.015	3.8	ug/L	823	Standard
	Co	59	2721.9	4.1	**0.3646**	0.019	5.1	ug/L	252	Standard
	Ni	60	2839.3	1.0	**1.4356**	0.026	1.8	ug/L	192	Standard
	Cu	65	1518.4	2.5	**0.7258**	0.025	3.5	ug/L	141	Standard
	Zn	66	6972.6	1.8	**5.6720**	0.065	1.1	ug/L	491	Standard
>	**Ge**	72	504116.3	0.9				ug/L	515087	Standard
	As	75	374.9	9.7	**0.3808**	0.031	8.0	ug/L	20	Standard
	Se	82	63.0	8.6	**0.4404**	0.052	11.9	ug/L	15	Standard
	Se-1	77	80.7	10.3	**0.3758**	0.122	32.5	ug/L	48	Standard
>	**Ga**	71	33.3	52.7				mg/L	90	Standard
	Rb	85	16.7	96.4				ug/L	23	Standard
	Y	89	427965.8	1.5				ug/L	448510	Standard
>	**Rh**	103	20.0	25.0				ug/L	15	Standard
	Mo	98	72.3	5.2	**0.0094**	0.001	12.5	ug/L	11	Standard
	Ag	107	2402.2	4.6	**0.3647**	0.015	4.1	ug/L	134	Standard
	Cd	111	440.5	3.4	**0.2165**	0.007	3.2	mg/L	7	Standard
	Cd	114	1158.4	0.8	**0.2324**	0.001	0.2	ug/L	16	Standard
>	**In**	115	574029.4	1.1				ug/L	611117	Standard
	Sn	118	496.7	16.4	**-0.0124**	0.016	125.3	ug/L	518	Standard
	Sb	123	1769.4	0.8	**0.3864**	0.003	0.7	ug/L	127	Standard
	Ba	135	1350.1	0.7	**0.6712**	0.012	1.8	ug/L	45	Standard
	Ce	140	15.0	0.0				ug/L	298	Standard
>	**Tb**	159	914394.0	1.1				ug/L	977942	Standard
	Ho	165	6.7	43.3				ug/L	12	Standard
	Tl	203	645.7	4.5	**0.0752**	0.004	4.8	ug/L	31	Standard
	Tl	205	546.7	2.9	**0.0757**	0.002	3.3	ug/L	20	Standard
	Pb	206	1234.7	1.4	**0.1748**	0.002	1.1	ug/L	286	Standard
	Pb	207	1077.0	1.2	**0.1725**	0.004	2.3	ug/L	257	Standard
	Pb	208	4484.6	3.2	**0.1877**	0.008	4.4	ug/L	956	Standard
	U	238	2602.6	1.7	**0.3761**	0.010	2.6	ug/L	17	Standard
>	**Bi**	209	533901.6	0.9				ug/L	557080	Standard

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 18:58:27
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	26.7	10.8	**-0.0196**	0.004	19.0	mg/L	32	Standard
K	39	28.3	10.2	**-0.0403**	0.010	24.0	mg/L	30	Standard
Ca	43	38.3	7.5	**0.8090**	0.513	63.4	mg/L	48	Standard
Fe	54	99.0	17.7	**-0.0423**	0.018	41.4	mg/L	159	Standard
Fe	57	198.3	16.4	**0.0831**	0.119	142.8	mg/L	195	Standard
Sc-1	45	37422.8	0.9				mg/L	39543	Standard
Cl	35	34311.5	2.9				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1263.4	4.0				ug/L	1647	Standard
P	31	12998.9	1.6				ug/L	15231	Standard
S	34	2533.5	5.7				ug/L	1845	Standard
Sr	88	71.7	29.0				ug/L	70	Standard
C	12	70.0	24.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	5.9	100.8				mg/L	12	Standard
Ho-1	165	6.7	43.3				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	1311.7	25.8				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	99.121		
Al	27			
Sc	45			
Ti	47			
V	51	77.290		
Cr	52	72.680		
Cr	53			
Mn	55	79.198		
Co	59	91.138		
Ni	60	89.723		
Cu	65	90.724		
Zn	66	90.752		
Ge	72		97.870	
As	75	95.209		
Se	82	110.096		
Se-1	77			
Ga	71			

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 18:58:27
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	91.178		
Cd	111	90.212		
Cd	114			
> In	115			93.931
Sn	118			
Sb	123	96.601		
Ba	135	89.490		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	93.978		
Tl	205			
Pb	206			
Pb	207			
Pb	208	93.825		
U	238	94.023		
> Bi	209			95.839
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 18:58:27
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: PBW A9 WG572242-02
Sample Date/Time: Tuesday, June 28, 2016 19:51:29
Number of Replicates: 3
Autosampler Position: 301
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73456.0	1.1				ug/L	81760	Standard
	Be	9	13.3	21.7	**0.0052**	0.003	59.5	ug/L	17	Standard
	Al	27	9322.9	1.9	**0.0529**	0.001	1.2	ug/L	1410	Standard
	Sc	45	38756.2	2.9				ug/L	39543	Standard
	Ti	47	33.7	15.2	**-0.0290**	0.022	74.1	ug/L	39	Standard
	V	51	1394.3	5.5	**-0.0051**	0.010	198.0	ug/L	1415	Standard
	Cr	52	7883.4	0.8	**0.0807**	0.013	16.5	ug/L	6946	Standard
	Cr	53	506.7	16.0	**0.2387**	0.095	39.9	ug/L	328	Standard
	Mn	55	1210.0	2.0	**0.0390**	0.005	11.7	ug/L	823	Standard
	Co	59	188.0	6.5	**-0.0045**	0.002	42.5	ug/L	252	Standard
	Ni	60	230.3	1.3	**0.0179**	0.002	10.0	ug/L	192	Standard
	Cu	65	157.0	8.9	**-0.0078**	0.008	102.0	ug/L	141	Standard
	Zn	66	1546.4	2.1	**0.5897**	0.023	3.8	ug/L	491	Standard
>	**Ge**	72	512703.5	0.7				ug/L	515087	Standard
	As	75	-32.6	48.2	**0.0111**	0.014	127.3	ug/L	20	Standard
	Se	82	19.2	20.1	**0.0486**	0.034	70.8	ug/L	15	Standard
	Se-1	77	50.0	10.6	**-0.0518**	0.075	145.3	ug/L	48	Standard
>	**Ga**	71	33.3	34.6				mg/L	90	Standard
	Rb	85	33.3	31.2				ug/L	23	Standard
	Y	89	438886.4	1.2				ug/L	448510	Standard
>	**Rh**	103	13.3	43.3				ug/L	15	Standard
	Mo	98	27.2	14.3	**-0.0088**	0.001	17.0	ug/L	11	Standard
	Ag	107	75.7	8.8	**-0.0047**	0.001	18.4	ug/L	134	Standard
	Cd	111	4.3	48.7	**-0.0009**	0.001	113.9	mg/L	7	Standard
	Cd	114	21.5	41.2	**-0.0007**	0.002	246.6	ug/L	16	Standard
>	**In**	115	598021.6	1.3				ug/L	611117	Standard
	Sn	118	688.3	6.3	**0.0170**	0.007	40.8	ug/L	518	Standard
	Sb	123	162.8	37.0	**0.0223**	0.013	56.3	ug/L	127	Standard
	Ba	135	50.0	12.0	**-0.0003**	0.003	1000.2	ug/L	45	Standard
	Ce	140	36.7	78.7				ug/L	298	Standard
>	**Tb**	159	940667.7	2.2				ug/L	977942	Standard
	Ho	165	18.3	56.8				ug/L	12	Standard
	Tl	203	9.0	22.2	**-0.0037**	0.000	6.2	ug/L	31	Standard
	Tl	205	5.0	0.0	**-0.0016**	0.000	0.7	ug/L	20	Standard
	Pb	206	302.0	8.3	**-0.0074**	0.006	75.2	ug/L	286	Standard
	Pb	207	237.7	5.4	**-0.0090**	0.003	32.3	ug/L	257	Standard
	Pb	208	966.3	1.0	**-0.0090**	0.001	6.5	ug/L	956	Standard
	U	238	4.3	48.0	**-0.0024**	0.000	12.3	ug/L	17	Standard
>	**Bi**	209	554162.6	1.6				ug/L	557080	Standard

Sample ID: PBW A9 WG572242-02
Report Date/Time: Tuesday, June 28, 2016 19:53:46
Page 1

Approved: June 29, 2016

	Na	23	0.0		**0.0050**			mg/L	0	Standard
	Mg	24	41.7	18.3	**-0.0005**	0.010	2017.0	mg/L	32	Standard
	K	39	30.0	33.3	**-0.0385**	0.033	84.7	mg/L	30	Standard
	Ca	43	16.7	91.7	**-2.7218**	2.434	89.4	mg/L	48	Standard
	Fe	54	121.1	16.6	**-0.0249**	0.021	84.8	mg/L	159	Standard
	Fe	57	230.0	15.2	**0.1658**	0.100	60.0	mg/L	195	Standard
>	**Sc-1**	45	38756.2	2.9				mg/L	39543	Standard
	Cl	35	36900.9	1.9				ug/L	27642	Standard
	Kr	83	1.3	43.3				ug/L	2	Standard
	Br	81	1656.8	8.0				ug/L	1647	Standard
	P	31	17782.1	2.2				ug/L	15231	Standard
	S	34	2408.5	10.1				ug/L	1845	Standard
	Sr	88	75.0	29.1				ug/L	70	Standard
	C	12	50.0	52.9				mg/L	113	Standard
	N	14	0.0					mg/L	0	Standard
	Hg	202	3.3	173.2				mg/L	0	Standard
	Dy	164	9.5	105.0				mg/L	12	Standard
	Ho-1	165	18.3	56.8				mg/L	12	Standard
	Er	166	10.0					mg/L	20	Standard
	I	127	1248.4	23.8				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		89.843	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		99.537	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBW A9 WG572242-02
Report Date/Time: Tuesday, June 28, 2016 19:53:46
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.857
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.476
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW A9 WG572242-02
Report Date/Time: Tuesday, June 28, 2016 19:53:46
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: LCSW A9 WG572242-03
Sample Date/Time: Tuesday, June 28, 2016 19:54:40
Number of Replicates: 3
Autosampler Position: 302
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74125.9	0.7				ug/L	81760	Standard
	Be	9	45613.6	2.5	**50.6567**	1.536	3.0	ug/L	17	Standard
	Al	27	52265.0	2.7	**0.3596**	0.008	2.3	ug/L	1410	Standard
	Sc	45	39738.7	2.1				ug/L	39543	Standard
	Ti	47	34.0	10.6	**-0.0277**	0.015	52.7	ug/L	39	Standard
	V	51	340531.5	1.6	**49.0319**	0.683	1.4	ug/L	1415	Standard
	Cr	52	354579.7	0.6	**50.1386**	0.494	1.0	ug/L	6946	Standard
	Cr	53	44436.8	3.3	**50.5389**	1.845	3.7	ug/L	328	Standard
	Mn	55	331885.6	0.6	**51.3034**	0.483	0.9	ug/L	823	Standard
	Co	59	341252.7	1.0	**48.7823**	0.497	1.0	ug/L	252	Standard
	Ni	60	94292.8	0.4	**50.2138**	0.089	0.2	ug/L	192	Standard
	Cu	65	96497.7	0.5	**50.9494**	0.130	0.3	ug/L	141	Standard
	Zn	66	57251.0	0.4	**51.6689**	0.417	0.8	ug/L	491	Standard
>	**Ge**	72	512516.5	0.4				ug/L	515087	Standard
	As	75	55356.1	0.6	**49.5225**	0.094	0.2	ug/L	20	Standard
	Se	82	5489.9	0.8	**47.8062**	0.574	1.2	ug/L	15	Standard
	Se-1	77	3699.5	2.3	**48.6318**	0.978	2.0	ug/L	48	Standard
>	**Ga**	71	51.7	29.6				mg/L	90	Standard
	Rb	85	58.3	27.6				ug/L	23	Standard
	Y	89	427399.3	0.3				ug/L	448510	Standard
>	**Rh**	103	20.0	66.1				ug/L	15	Standard
	Mo	98	50.9	5.4	**0.0000**	0.001	4406.4	ug/L	11	Standard
	Ag	107	325451.5	0.2	**49.1242**	0.412	0.8	ug/L	134	Standard
	Cd	111	100950.5	1.5	**47.8617**	0.611	1.3	mg/L	7	Standard
	Cd	114	243935.8	0.5	**47.5806**	0.389	0.8	ug/L	16	Standard
>	**In**	115	602979.8	0.8				ug/L	611117	Standard
	Sn	118	653.3	9.2	**0.0100**	0.010	97.1	ug/L	518	Standard
	Sb	123	224336.4	0.4	**48.1778**	0.461	1.0	ug/L	127	Standard
	Ba	135	98286.4	0.2	**48.2270**	0.385	0.8	ug/L	45	Standard
	Ce	140	205.0	19.4				ug/L	298	Standard
>	**Tb**	159	962904.7	0.9				ug/L	977942	Standard
	Ho	165	11.7	49.5				ug/L	12	Standard
	Tl	203	409811.5	0.9	**48.6592**	0.503	1.0	ug/L	31	Standard
	Tl	205	358097.5	0.9	**48.9662**	0.679	1.4	ug/L	20	Standard
	Pb	206	272798.6	0.1	**50.4414**	0.712	1.4	ug/L	286	Standard
	Pb	207	236678.2	0.4	**48.5210**	0.945	1.9	ug/L	257	Standard
	Pb	208	941837.1	1.1	**49.9011**	0.393	0.8	ug/L	956	Standard
	U	238	343975.4	0.3	**48.0328**	0.703	1.5	ug/L	17	Standard
>	**Bi**	209	556837.1	1.5				ug/L	557080	Standard

Sample ID: LCSW A9 WG572242-03
Report Date/Time: Tuesday, June 28, 2016 19:56:57
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.3983**	18.002	173.1	mg/L	0	Standard
Mg	24	58.3	27.6	**0.0205**	0.023	110.8	mg/L	32	Standard
K	39	16.7	45.8	**-0.0851**	0.027	31.2	mg/L	30	Standard
Ca	43	23.3	32.7	**-1.8206**	1.160	63.7	mg/L	48	Standard
Fe	54	118.7	10.7	**-0.0303**	0.010	31.4	mg/L	159	Standard
Fe	57	200.0	10.0	**0.0482**	0.079	163.5	mg/L	195	Standard
Sc-1	45	39738.7	2.1				mg/L	39543	Standard
Cl	35	36016.8	1.9				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	3627.1	7.0				ug/L	1647	Standard
P	31	18584.7	0.6				ug/L	15231	Standard
S	34	2266.8	5.1				ug/L	1845	Standard
Sr	88	53.3	23.6				ug/L	70	Standard
C	12	126.7	32.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	0	Standard
Dy	164	9.0	105.3				mg/L	12	Standard
Ho-1	165	11.7	49.5				mg/L	12	Standard
Er	166	20.0	50.0				mg/L	20	Standard
I	127	1253.4	17.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		90.663	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.501	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW A9 WG572242-03
Report Date/Time: Tuesday, June 28, 2016 19:56:57
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.669
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.956
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW A9 WG572242-03
Report Date/Time: Tuesday, June 28, 2016 19:56:57
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606083901 WG574242-01
Sample Date/Time: Tuesday, June 28, 2016 19:57:52
Number of Replicates: 3
Autosampler Position: 303
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73849.6	1.9				ug/L	81760	Standard
	Be	9	25.0	0.0	**0.0181**	0.001	2.9	ug/L	17	Standard
	Al	27	15521751.4	1.0	**111.4854**	1.467	1.3	ug/L	1410	Standard
	Sc	45	39103.8	3.8				ug/L	39543	Standard
	Ti	47	591.7	5.1	**2.3246**	0.153	6.6	ug/L	39	Standard
	V	51	4219.1	3.4	**0.4282**	0.015	3.5	ug/L	1415	Standard
	Cr	52	8895.6	2.0	**0.2798**	0.040	14.2	ug/L	6946	Standard
	Cr	53	3360.4	10.4	**3.6642**	0.437	11.9	ug/L	328	Standard
	Mn	55	16077.2	1.4	**2.4454**	0.007	0.3	ug/L	823	Standard
	Co	59	1032.0	3.9	**0.1223**	0.006	4.9	ug/L	252	Standard
	Ni	60	3143.3	2.0	**1.6407**	0.018	1.1	ug/L	192	Standard
	Cu	65	3355.0	2.0	**1.7559**	0.016	0.9	ug/L	141	Standard
	Zn	66	14991.7	0.7	**13.4797**	0.137	1.0	ug/L	491	Standard
>	**Ge**	72	492452.2	1.1				ug/L	515087	Standard
	As	75	562.8	15.9	**0.5640**	0.084	14.9	ug/L	20	Standard
	Se	82	67.0	15.0	**0.4895**	0.090	18.3	ug/L	15	Standard
	Se-1	77	207.3	11.8	**2.1604**	0.344	15.9	ug/L	48	Standard
>	**Ga**	71	103.3	23.9				mg/L	90	Standard
	Rb	85	20340.3	1.9				ug/L	23	Standard
	Y	89	439150.3	1.0				ug/L	448510	Standard
>	**Rh**	103	18.3	15.7				ug/L	15	Standard
	Mo	98	1185.2	3.1	**0.4267**	0.016	3.7	ug/L	11	Standard
	Ag	107	124.0	3.2	**0.0026**	0.001	26.5	ug/L	134	Standard
	Cd	111	35.9	29.2	**0.0141**	0.005	34.6	mg/L	7	Standard
	Cd	114	96.0	6.6	**0.0139**	0.001	9.0	ug/L	16	Standard
>	**In**	115	601194.2	0.6				ug/L	611117	Standard
	Sn	118	770.0	3.2	**0.0304**	0.004	13.3	ug/L	518	Standard
	Sb	123	1910.7	20.7	**0.3992**	0.088	22.0	ug/L	127	Standard
	Ba	135	33674.7	0.4	**16.5556**	0.054	0.3	ug/L	45	Standard
	Ce	140	1643.4	3.5				ug/L	298	Standard
>	**Tb**	159	946228.9	1.7				ug/L	977942	Standard
	Ho	165	41.7	36.7				ug/L	12	Standard
	Tl	203	211.3	8.6	**0.0215**	0.002	11.4	ug/L	31	Standard
	Tl	205	188.3	11.1	**0.0247**	0.003	12.3	ug/L	20	Standard
	Pb	206	2721.9	1.8	**0.4638**	0.014	3.1	ug/L	286	Standard
	Pb	207	2208.8	1.5	**0.4164**	0.010	2.4	ug/L	257	Standard
	Pb	208	9386.5	1.8	**0.4607**	0.014	3.1	ug/L	956	Standard
	U	238	2364.2	1.6	**0.3425**	0.008	2.2	ug/L	17	Standard
>	**Bi**	209	532104.5	1.0				ug/L	557080	Standard

Sample ID: L1606083901 WG574242-01
Report Date/Time: Tuesday, June 28, 2016 20:00:08
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.9240**	18.912	173.1	mg/L	0	Standard
Mg	24	14932.4	1.8	**20.0717**	1.109	5.5	mg/L	32	Standard
K	39	851.7	12.5	**2.7753**	0.419	15.1	mg/L	30	Standard
Ca	43	323.3	7.9	**44.3168**	3.500	7.9	mg/L	48	Standard
Fe	54	180.7	15.7	**0.0282**	0.024	84.5	mg/L	159	Standard
Fe	57	355.0	6.1	**0.5869**	0.046	7.9	mg/L	195	Standard
> **Sc-1**	45	39103.8	3.8				mg/L	39543	Standard
Cl	35	38446.7	1.8				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	10497.0	5.2				ug/L	1647	Standard
P	31	21567.0	1.7				ug/L	15231	Standard
S	34	2493.5	2.8				ug/L	1845	Standard
Sr	88	151.7	7.6				ug/L	70	Standard
C	12	210.0	39.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	67.9	17.5				mg/L	12	Standard
Ho-1	165	41.7	36.7				mg/L	12	Standard
Er	166	43.3	93.3				mg/L	20	Standard
I	127	29340.1	3.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		90.325	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.606	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606083901 WG574242-01
Report Date/Time: Tuesday, June 28, 2016 20:00:08
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.376
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.517
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606083901DP WG574242-04
Sample Date/Time: Tuesday, June 28, 2016 20:01:03
Number of Replicates: 3
Autosampler Position: 304
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74496.0	1.1				ug/L	81760	Standard
	Be	9	20.0	25.0	**0.0124**	0.006	45.8	ug/L	17	Standard
	Al	27	15458448.5	1.4	**110.0484**	0.587	0.5	ug/L	1410	Standard
	Sc	45	38253.3	3.4				ug/L	39543	Standard
	Ti	47	558.7	5.4	**2.1548**	0.155	7.2	ug/L	39	Standard
	V	51	4067.8	2.6	**0.3973**	0.009	2.2	ug/L	1415	Standard
	Cr	52	9024.7	1.6	**0.2811**	0.031	11.1	ug/L	6946	Standard
	Cr	53	3592.1	3.4	**3.8809**	0.090	2.3	ug/L	328	Standard
	Mn	55	13517.4	2.1	**2.0039**	0.068	3.4	ug/L	823	Standard
	Co	59	1022.7	3.5	**0.1189**	0.006	5.0	ug/L	252	Standard
	Ni	60	3054.6	1.9	**1.5691**	0.041	2.6	ug/L	192	Standard
	Cu	65	3452.1	3.2	**1.7838**	0.034	1.9	ug/L	141	Standard
	Zn	66	14844.3	0.5	**13.1506**	0.169	1.3	ug/L	491	Standard
>	**Ge**	72	499109.4	1.4				ug/L	515087	Standard
	As	75	584.5	8.0	**0.5768**	0.041	7.2	ug/L	20	Standard
	Se	82	58.0	10.0	**0.3998**	0.046	11.4	ug/L	15	Standard
	Se-1	77	200.7	10.2	**2.0328**	0.318	15.6	ug/L	48	Standard
>	**Ga**	71	100.0	13.2				mg/L	90	Standard
	Rb	85	19739.5	1.7				ug/L	23	Standard
	Y	89	436250.9	1.8				ug/L	448510	Standard
>	**Rh**	103	38.3	7.5				ug/L	15	Standard
	Mo	98	1219.4	1.0	**0.4525**	0.003	0.7	ug/L	11	Standard
	Ag	107	99.7	5.7	**-0.0007**	0.001	125.7	ug/L	134	Standard
	Cd	111	32.6	12.4	**0.0130**	0.002	14.9	mg/L	7	Standard
	Cd	114	81.5	18.5	**0.0115**	0.003	24.6	ug/L	16	Standard
>	**In**	115	584572.0	1.4				ug/L	611117	Standard
	Sn	118	751.7	6.1	**0.0310**	0.009	27.5	ug/L	518	Standard
	Sb	123	786.1	12.1	**0.1616**	0.023	14.4	ug/L	127	Standard
	Ba	135	32184.1	0.9	**16.2729**	0.095	0.6	ug/L	45	Standard
	Ce	140	490.0	7.4				ug/L	298	Standard
>	**Tb**	159	929479.8	0.3				ug/L	977942	Standard
	Ho	165	31.7	24.1				ug/L	12	Standard
	Tl	203	152.7	10.6	**0.0145**	0.002	14.7	ug/L	31	Standard
	Tl	205	146.7	17.2	**0.0191**	0.004	18.8	ug/L	20	Standard
	Pb	206	2610.2	1.4	**0.4511**	0.005	1.2	ug/L	286	Standard
	Pb	207	2110.5	2.5	**0.4033**	0.013	3.2	ug/L	257	Standard
	Pb	208	8866.0	0.9	**0.4405**	0.006	1.4	ug/L	956	Standard
	U	238	2335.8	3.1	**0.3444**	0.010	2.9	ug/L	17	Standard
>	**Bi**	209	522741.7	0.4				ug/L	557080	Standard

Sample ID: L1606083901DP WG574242-04
Report Date/Time: Tuesday, June 28, 2016 20:03:20
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**11.1196**	19.251	173.1	mg/L	0	Standard
Mg	24	14720.5	1.4	**20.2196**	0.936	4.6	mg/L	32	Standard
K	39	903.4	10.8	**3.0129**	0.317	10.5	mg/L	30	Standard
Ca	43	308.3	1.9	**43.1033**	1.939	4.5	mg/L	48	Standard
Fe	54	132.3	23.0	**-0.0139**	0.024	174.3	mg/L	159	Standard
Fe	57	371.7	11.8	**0.6755**	0.185	27.4	mg/L	195	Standard
> **Sc-1**	45	38253.3	3.4				mg/L	39543	Standard
Cl	35	39853.0	2.0				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	7995.4	3.8				ug/L	1647	Standard
P	31	20689.1	2.4				ug/L	15231	Standard
S	34	2350.2	5.7				ug/L	1845	Standard
Sr	88	115.0	7.5				ug/L	70	Standard
C	12	266.7	11.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	41.7	57.7				mg/L	12	Standard
Ho-1	165	31.7	24.1				mg/L	12	Standard
Er	166	33.3	75.5				mg/L	20	Standard
I	127	28308.1	3.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.115	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.898	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606083901DP WG574242-04
Report Date/Time: Tuesday, June 28, 2016 20:03:20
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.656
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		93.836
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606083901S WG574242-05
Sample Date/Time: Tuesday, June 28, 2016 20:04:14
Number of Replicates: 3
Autosampler Position: 305
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77052.1	1.2				ug/L	81760	Standard
	Be	9	48365.4	1.4	**51.6767**	1.282	2.5	ug/L	17	Standard
	Al	27	15205678.0	0.8	**104.6750**	1.999	1.9	ug/L	1410	Standard
	Sc	45	39767.1	2.3				ug/L	39543	Standard
	Ti	47	566.0	2.6	**2.1182**	0.041	1.9	ug/L	39	Standard
	V	51	345428.5	0.5	**49.6572**	0.625	1.3	ug/L	1415	Standard
	Cr	52	342558.9	0.4	**48.3219**	0.886	1.8	ug/L	6946	Standard
	Cr	53	45610.2	2.6	**51.7999**	1.901	3.7	ug/L	328	Standard
	Mn	55	334844.2	1.7	**51.6671**	0.130	0.3	ug/L	823	Standard
	Co	59	340587.2	2.1	**48.5963**	0.256	0.5	ug/L	252	Standard
	Ni	60	92425.4	0.6	**49.1352**	0.675	1.4	ug/L	192	Standard
	Cu	65	94140.2	0.6	**49.6217**	0.865	1.7	ug/L	141	Standard
	Zn	66	68169.3	0.4	**61.5791**	1.173	1.9	ug/L	491	Standard
>	**Ge**	72	513441.7	1.6				ug/L	515087	Standard
	As	75	56899.2	0.6	**50.8164**	0.583	1.1	ug/L	20	Standard
	Se	82	5653.4	1.3	**49.1454**	0.462	0.9	ug/L	15	Standard
	Se-1	77	3932.8	1.5	**51.6632**	1.250	2.4	ug/L	48	Standard
>	**Ga**	71	93.3	35.7				mg/L	90	Standard
	Rb	85	19300.6	2.7				ug/L	23	Standard
	Y	89	425453.4	2.7				ug/L	448510	Standard
>	**Rh**	103	45.0	19.2				ug/L	15	Standard
	Mo	98	1218.6	0.9	**0.4503**	0.008	1.7	ug/L	11	Standard
	Ag	107	314898.7	0.6	**48.8175**	0.102	0.2	ug/L	134	Standard
	Cd	111	99777.7	0.5	**48.5906**	0.622	1.3	mg/L	7	Standard
	Cd	114	243059.8	0.9	**48.6969**	0.795	1.6	ug/L	16	Standard
>	**In**	115	587073.7	0.8				ug/L	611117	Standard
	Sn	118	733.4	7.3	**0.0272**	0.010	36.7	ug/L	518	Standard
	Sb	123	223776.0	1.4	**49.3561**	0.290	0.6	ug/L	127	Standard
	Ba	135	127336.9	0.7	**64.1810**	0.344	0.5	ug/L	45	Standard
	Ce	140	335.0	9.1				ug/L	298	Standard
>	**Tb**	159	937062.8	1.0				ug/L	977942	Standard
	Ho	165	41.7	79.9				ug/L	12	Standard
	Tl	203	401801.7	0.6	**49.2378**	0.579	1.2	ug/L	31	Standard
	Tl	205	351031.6	0.7	**49.5383**	0.691	1.4	ug/L	20	Standard
	Pb	206	267916.7	0.7	**51.1246**	0.591	1.2	ug/L	286	Standard
	Pb	207	234005.3	0.9	**49.5047**	0.372	0.8	ug/L	257	Standard
	Pb	208	924903.3	0.8	**50.5781**	0.765	1.5	ug/L	956	Standard
	U	238	349197.5	0.5	**50.3229**	0.614	1.2	ug/L	17	Standard
>	**Bi**	209	539520.7	0.7				ug/L	557080	Standard

Sample ID: L1606083901S WG574242-05
Report Date/Time: Tuesday, June 28, 2016 20:06:31
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.8832**	18.842	173.1	mg/L	0	Standard
Mg	24	14542.0	3.3	**19.1893**	0.286	1.5	mg/L	32	Standard
K	39	753.4	10.6	**2.3925**	0.305	12.7	mg/L	30	Standard
Ca	43	295.0	12.2	**39.3051**	6.421	16.3	mg/L	48	Standard
Fe	54	140.6	16.7	**-0.0107**	0.020	183.5	mg/L	159	Standard
Fe	57	348.3	9.1	**0.5437**	0.082	15.1	mg/L	195	Standard
> **Sc-1**	45	39767.1	2.3				mg/L	39543	Standard
Cl	35	39860.4	1.2				ug/L	27642	Standard
Kr	83	2.3	49.5				ug/L	2	Standard
Br	81	7565.2	10.7				ug/L	1647	Standard
P	31	21044.6	3.8				ug/L	15231	Standard
S	34	2540.2	5.7				ug/L	1845	Standard
Sr	88	145.0	22.6				ug/L	70	Standard
C	12	196.7	11.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	48.4	54.7				mg/L	12	Standard
Ho-1	165	41.7	79.9				mg/L	12	Standard
Er	166	33.3	45.8				mg/L	20	Standard
I	127	26943.9	1.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.242	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.681	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606083901S WG574242-05
Report Date/Time: Tuesday, June 28, 2016 20:06:31
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.066
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.848
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606083901S WG574242-05
Report Date/Time: Tuesday, June 28, 2016 20:06:31
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606083901SD WG574242-06
Sample Date/Time: Tuesday, June 28, 2016 20:07:26
Number of Replicates: 3
Autosampler Position: 306
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77184.4	1.1				ug/L	81760	Standard
	Be	9	48315.3	1.2	**51.5258**	0.436	0.8	ug/L	17	Standard
	Al	27	15387465.4	1.3	**105.7299**	0.773	0.7	ug/L	1410	Standard
	Sc	45	40039.5	2.3				ug/L	39543	Standard
	Ti	47	554.0	4.1	**2.1409**	0.093	4.3	ug/L	39	Standard
	V	51	332776.2	1.5	**49.3498**	0.268	0.5	ug/L	1415	Standard
	Cr	52	340776.8	0.8	**49.6175**	0.095	0.2	ug/L	6946	Standard
	Cr	53	45466.5	0.7	**53.2763**	0.775	1.5	ug/L	328	Standard
	Mn	55	324192.4	1.4	**51.6123**	0.258	0.5	ug/L	823	Standard
	Co	59	327045.3	2.1	**48.1461**	0.569	1.2	ug/L	252	Standard
	Ni	60	91364.7	1.1	**50.1099**	0.072	0.1	ug/L	192	Standard
	Cu	65	91854.3	0.9	**49.9479**	0.081	0.2	ug/L	141	Standard
	Zn	66	66866.6	0.8	**62.3204**	0.118	0.2	ug/L	491	Standard
>	**Ge**	72	497622.9	1.0				ug/L	515087	Standard
	As	75	54845.8	1.7	**50.5323**	0.480	1.0	ug/L	20	Standard
	Se	82	5429.5	1.2	**48.6954**	0.265	0.5	ug/L	15	Standard
	Se-1	77	3850.8	1.9	**52.1872**	0.544	1.0	ug/L	48	Standard
>	**Ga**	71	85.0	11.8				mg/L	90	Standard
	Rb	85	19015.3	1.2				ug/L	23	Standard
	Y	89	421533.6	0.8				ug/L	448510	Standard
>	**Rh**	103	55.0	15.7				ug/L	15	Standard
	Mo	98	1205.1	2.4	**0.4408**	0.012	2.7	ug/L	11	Standard
	Ag	107	307002.5	0.8	**47.1551**	0.386	0.8	ug/L	134	Standard
	Cd	111	97425.8	0.9	**47.0062**	0.430	0.9	mg/L	7	Standard
	Cd	114	239795.9	0.5	**47.5975**	0.300	0.6	ug/L	16	Standard
>	**In**	115	592518.8	0.1				ug/L	611117	Standard
	Sn	118	761.7	9.1	**0.0309**	0.012	38.7	ug/L	518	Standard
	Sb	123	220298.1	1.2	**48.1440**	0.631	1.3	ug/L	127	Standard
	Ba	135	125301.0	0.6	**62.5732**	0.412	0.7	ug/L	45	Standard
	Ce	140	340.0	1.5				ug/L	298	Standard
>	**Tb**	159	948091.3	1.9				ug/L	977942	Standard
	Ho	165	23.3	65.5				ug/L	12	Standard
	Tl	203	394026.8	0.4	**48.4831**	0.151	0.3	ug/L	31	Standard
	Tl	205	348699.1	1.4	**49.4146**	1.048	2.1	ug/L	20	Standard
	Pb	206	261711.9	1.4	**50.1430**	0.483	1.0	ug/L	286	Standard
	Pb	207	228078.7	1.5	**48.4479**	0.479	1.0	ug/L	257	Standard
	Pb	208	906176.6	0.5	**49.7571**	0.517	1.0	ug/L	956	Standard
	U	238	340376.9	0.5	**49.2541**	0.514	1.0	ug/L	17	Standard
>	**Bi**	209	537295.6	0.7				ug/L	557080	Standard

Sample ID: L1606083901SD WG574242-06
Report Date/Time: Tuesday, June 28, 2016 20:09:43
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**20.8666**	18.067	86.6	mg/L	0	Standard
Mg	24	14582.0	2.3	**19.1163**	0.419	2.2	mg/L	32	Standard
K	39	795.0	3.8	**2.5139**	0.164	6.5	mg/L	30	Standard
Ca	43	345.0	2.9	**46.4459**	2.697	5.8	mg/L	48	Standard
Fe	54	142.5	14.7	**-0.0097**	0.018	187.3	mg/L	159	Standard
Fe	57	408.3	2.5	**0.7378**	0.065	8.9	mg/L	195	Standard
> **Sc-1**	45	40039.5	2.3				mg/L	39543	Standard
Cl	35	39300.2	1.1				ug/L	27642	Standard
Kr	83	1.7	173.2				ug/L	2	Standard
Br	81	7651.9	3.5				ug/L	1647	Standard
P	31	20634.0	3.2				ug/L	15231	Standard
S	34	2398.5	2.0				ug/L	1845	Standard
Sr	88	180.0	15.5				ug/L	70	Standard
C	12	226.7	27.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	27.4	58.7				mg/L	12	Standard
Ho-1	165	23.3	65.5				mg/L	12	Standard
Er	166	53.3	108.3				mg/L	20	Standard
I	127	27065.8	3.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.404	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.610	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606083901SD WG574242-06
Report Date/Time: Tuesday, June 28, 2016 20:09:43
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.957
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.448
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606083901SD WG574242-06
Report Date/Time: Tuesday, June 28, 2016 20:09:43
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606084301
Sample Date/Time: Tuesday, June 28, 2016 20:10:37
Number of Replicates: 3
Autosampler Position: 307
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80372.6	1.6				ug/L	81760	Standard
	Be	9	41.7	45.4	**0.0330**	0.020	59.2	ug/L	17	Standard
	Al	27	7042054.5	1.4	**46.4726**	1.334	2.9	ug/L	1410	Standard
	Sc	45	38211.6	5.6				ug/L	39543	Standard
	Ti	47	149.3	19.8	**0.4676**	0.133	28.4	ug/L	39	Standard
	V	51	1769.8	3.9	**0.0612**	0.007	10.9	ug/L	1415	Standard
	Cr	52	8360.6	4.6	**0.2060**	0.044	21.1	ug/L	6946	Standard
	Cr	53	938.4	5.4	**0.7826**	0.048	6.1	ug/L	328	Standard
	Mn	55	4412856.2	0.9	**715.8990**	5.361	0.7	ug/L	823	Standard
	Co	59	49628.8	1.9	**7.3983**	0.034	0.5	ug/L	252	Standard
	Ni	60	31518.1	1.1	**17.4994**	0.073	0.4	ug/L	192	Standard
	Cu	65	1083.7	2.3	**0.5092**	0.018	3.5	ug/L	141	Standard
	Zn	66	58737.0	1.3	**55.5464**	0.787	1.4	ug/L	491	Standard
>	**Ge**	72	489683.3	1.5				ug/L	515087	Standard
	As	75	118.1	28.8	**0.1508**	0.032	21.2	ug/L	20	Standard
	Se	82	133.7	3.3	**1.1023**	0.049	4.4	ug/L	15	Standard
	Se-1	77	146.3	10.3	**1.3266**	0.239	18.0	ug/L	48	Standard
>	**Ga**	71	230.0	39.9				mg/L	90	Standard
	Rb	85	13262.5	1.5				ug/L	23	Standard
	Y	89	437622.3	1.3				ug/L	448510	Standard
>	**Rh**	103	40.0	43.3				ug/L	15	Standard
	Mo	98	84.3	2.5	**0.0124**	0.001	5.4	ug/L	11	Standard
	Ag	107	135.7	14.4	**0.0042**	0.003	69.6	ug/L	134	Standard
	Cd	111	325.5	6.3	**0.1508**	0.009	6.1	mg/L	7	Standard
	Cd	114	800.8	5.2	**0.1507**	0.008	5.1	ug/L	16	Standard
>	**In**	115	605376.8	0.3				ug/L	611117	Standard
	Sn	118	716.7	15.8	**0.0204**	0.019	93.4	ug/L	518	Standard
	Sb	123	1846.5	24.8	**0.3824**	0.099	25.9	ug/L	127	Standard
	Ba	135	61196.0	0.7	**29.8977**	0.107	0.4	ug/L	45	Standard
	Ce	140	4327.3	3.4				ug/L	298	Standard
>	**Tb**	159	946751.9	1.1				ug/L	977942	Standard
	Ho	165	113.3	17.8				ug/L	12	Standard
	Tl	203	274.7	6.5	**0.0286**	0.002	7.3	ug/L	31	Standard
	Tl	205	308.3	53.1	**0.0408**	0.023	55.8	ug/L	20	Standard
	Pb	206	584.3	5.1	**0.0470**	0.006	11.8	ug/L	286	Standard
	Pb	207	484.3	1.7	**0.0436**	0.002	3.9	ug/L	257	Standard
	Pb	208	2242.1	13.2	**0.0611**	0.016	25.8	ug/L	956	Standard
	U	238	167.0	68.8	**0.0208**	0.016	78.4	ug/L	17	Standard
>	**Bi**	209	544741.6	0.3				ug/L	557080	Standard

Sample ID: L1606084301
Report Date/Time: Tuesday, June 28, 2016 20:12:54
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	6142.9	1.9	**8.4190**	0.399	4.7	mg/L	32	Standard
K	39	320.0	19.2	**0.9720**	0.149	15.3	mg/L	30	Standard
Ca	43	193.3	18.7	**25.2140**	6.755	26.8	mg/L	48	Standard
Fe	54	272.9	8.2	**0.1197**	0.034	28.2	mg/L	159	Standard
Fe	57	298.3	20.0	**0.4171**	0.202	48.3	mg/L	195	Standard
> **Sc-1**	45	38211.6	5.6				mg/L	39543	Standard
Cl	35	43040.4	2.1				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	3493.7	8.6				ug/L	1647	Standard
P	31	19535.9	3.1				ug/L	15231	Standard
S	34	2298.5	3.2				ug/L	1845	Standard
Sr	88	191.7	4.0				ug/L	70	Standard
C	12	103.3	24.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	171.7	2.4				mg/L	12	Standard
Ho-1	165	113.3	17.8				mg/L	12	Standard
Er	166	103.3	43.6				mg/L	20	Standard
I	127	15184.3	1.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		98.303	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.068	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606084301
Report Date/Time: Tuesday, June 28, 2016 20:12:54
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.061
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.785
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606084301
Report Date/Time: Tuesday, June 28, 2016 20:12:54
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606084301PS WG574570-01
Sample Date/Time: Tuesday, June 28, 2016 20:13:49
Number of Replicates: 3
Autosampler Position: 308
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84132.9	0.3				ug/L	81760	Standard
	Be	9	53800.4	2.5	**52.6389**	1.459	2.8	ug/L	17	Standard
	Al	27	7286262.3	0.5	**45.9217**	0.095	0.2	ug/L	1410	Standard
	Sc	45	39187.3	1.8				ug/L	39543	Standard
	Ti	47	167.0	8.4	**0.5053**	0.064	12.8	ug/L	39	Standard
	V	51	355576.0	1.4	**50.8299**	0.802	1.6	ug/L	1415	Standard
	Cr	52	356673.6	1.0	**50.0684**	1.233	2.5	ug/L	6946	Standard
	Cr	53	44503.6	1.2	**50.2498**	1.575	3.1	ug/L	328	Standard
	Mn	55	4859292.5	1.8	**747.6759**	17.580	2.4	ug/L	823	Standard
	Co	59	417615.7	1.3	**59.2660**	0.892	1.5	ug/L	252	Standard
	Ni	60	125871.4	1.4	**66.5870**	2.033	3.1	ug/L	192	Standard
	Cu	65	97106.5	1.0	**50.8947**	0.717	1.4	ug/L	141	Standard
	Zn	66	127581.2	0.7	**115.3025**	1.913	1.7	ug/L	491	Standard
>	**Ge**	72	516403.0	2.2				ug/L	515087	Standard
	As	75	57691.6	2.0	**51.2302**	0.986	1.9	ug/L	20	Standard
	Se	82	6053.8	2.8	**52.3316**	1.116	2.1	ug/L	15	Standard
	Se-1	77	3949.2	3.8	**51.5591**	1.119	2.2	ug/L	48	Standard
>	**Ga**	71	251.7	10.0				mg/L	90	Standard
	Rb	85	13841.3	3.1				ug/L	23	Standard
	Y	89	444794.9	3.0				ug/L	448510	Standard
>	**Rh**	103	78.3	24.2				ug/L	15	Standard
	Mo	98	86.4	4.3	**0.0128**	0.001	7.7	ug/L	11	Standard
	Ag	107	327259.4	2.3	**48.6313**	0.643	1.3	ug/L	134	Standard
	Cd	111	105349.6	1.3	**49.1793**	0.322	0.7	mg/L	7	Standard
	Cd	114	257271.1	1.0	**49.4173**	1.101	2.2	ug/L	16	Standard
>	**In**	115	612449.2	2.0				ug/L	611117	Standard
	Sn	118	941.7	3.9	**0.0570**	0.004	7.5	ug/L	518	Standard
	Sb	123	235069.6	1.3	**49.7050**	0.359	0.7	ug/L	127	Standard
	Ba	135	161643.1	1.4	**78.1070**	0.551	0.7	ug/L	45	Standard
	Ce	140	4429.0	1.8				ug/L	298	Standard
>	**Tb**	159	958659.2	2.3				ug/L	977942	Standard
	Ho	165	101.7	15.0				ug/L	12	Standard
	Tl	203	432017.4	2.0	**50.9163**	0.725	1.4	ug/L	31	Standard
	Tl	205	384793.0	2.1	**52.2411**	1.764	3.4	ug/L	20	Standard
	Pb	206	286141.9	1.8	**52.5195**	0.925	1.8	ug/L	286	Standard
	Pb	207	249240.4	1.8	**50.7143**	0.544	1.1	ug/L	257	Standard
	Pb	208	995726.4	2.1	**52.3723**	1.044	2.0	ug/L	956	Standard
	U	238	370701.4	1.5	**51.3906**	1.394	2.7	ug/L	17	Standard
>	**Bi**	209	561025.3	2.5				ug/L	557080	Standard

Sample ID: L1606084301PS WG574570-01
Report Date/Time: Tuesday, June 28, 2016 20:16:06
Page 1

Approved: June 29, 2016

Na	23	5.0	0.0	**32.4034**	0.594	1.8	mg/L	0	Standard
Mg	24	6414.7	1.1	**8.5623**	0.246	2.9	mg/L	32	Standard
K	39	321.7	9.1	**0.9561**	0.110	11.5	mg/L	30	Standard
Ca	43	185.0	8.1	**22.9936**	1.940	8.4	mg/L	48	Standard
Fe	54	250.2	14.1	**0.0912**	0.029	32.3	mg/L	159	Standard
Fe	57	306.7	15.8	**0.4183**	0.145	34.7	mg/L	195	Standard
> **Sc-1**	45	39187.3	1.8				mg/L	39543	Standard
Cl	35	43079.8	1.6				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	3410.4	11.2				ug/L	1647	Standard
P	31	19879.7	4.3				ug/L	15231	Standard
S	34	2313.5	2.6				ug/L	1845	Standard
Sr	88	201.7	5.2				ug/L	70	Standard
C	12	140.0	32.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0					mg/L	0	Standard
Dy	164	196.2	20.7				mg/L	12	Standard
Ho-1	165	101.7	15.0				mg/L	12	Standard
Er	166	80.0	76.0				mg/L	20	Standard
I	127	16267.1	4.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.902	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.256	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606084301PS WG574570-01
Report Date/Time: Tuesday, June 28, 2016 20:16:06
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.218
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.708
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606084301PS WG574570-01
Report Date/Time: Tuesday, June 28, 2016 20:16:06
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606084301SDL WG574570-02
Sample Date/Time: Tuesday, June 28, 2016 20:17:00
Number of Replicates: 3
Autosampler Position: 309
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84610.6	1.2				ug/L	81760	Standard
	Be	9	31.7	24.1	**0.0210**	0.007	33.6	ug/L	17	Standard
	Al	27	1535529.0	0.5	**9.6123**	0.084	0.9	ug/L	1410	Standard
	Sc	45	40586.0	2.0				ug/L	39543	Standard
	Ti	47	54.0	14.5	**0.0442**	0.029	66.4	ug/L	39	Standard
	V	51	1510.6	4.6	**0.0030**	0.011	356.0	ug/L	1415	Standard
	Cr	52	7700.6	1.7	**0.0097**	0.022	221.3	ug/L	6946	Standard
	Cr	53	548.3	6.2	**0.2609**	0.037	14.0	ug/L	328	Standard
	Mn	55	927997.3	1.4	**137.8897**	1.374	1.0	ug/L	823	Standard
	Co	59	10741.5	3.1	**1.4429**	0.042	2.9	ug/L	252	Standard
	Ni	60	6856.2	1.9	**3.4057**	0.060	1.7	ug/L	192	Standard
	Cu	65	384.3	5.5	**0.1042**	0.011	10.3	ug/L	141	Standard
	Zn	66	13504.0	1.8	**11.0531**	0.162	1.5	ug/L	491	Standard
>	**Ge**	72	534132.0	0.6				ug/L	515087	Standard
	As	75	-8.2	426.8	**0.0333**	0.030	90.7	ug/L	20	Standard
	Se	82	34.3	14.2	**0.1678**	0.041	24.5	ug/L	15	Standard
	Se-1	77	77.7	7.3	**0.2756**	0.079	28.6	ug/L	48	Standard
>	**Ga**	71	133.3	12.1				mg/L	90	Standard
	Rb	85	2885.3	6.5				ug/L	23	Standard
	Y	89	438369.8	1.0				ug/L	448510	Standard
>	**Rh**	103	21.7	13.3				ug/L	15	Standard
	Mo	98	33.0	37.7	**-0.0070**	0.005	64.6	ug/L	11	Standard
	Ag	107	108.0	7.3	**-0.0003**	0.002	523.1	ug/L	134	Standard
	Cd	111	82.3	4.3	**0.0351**	0.001	2.2	mg/L	7	Standard
	Cd	114	247.3	23.3	**0.0421**	0.011	25.9	ug/L	16	Standard
>	**In**	115	618448.1	2.5				ug/L	611117	Standard
	Sn	118	536.7	11.9	**-0.0124**	0.008	66.9	ug/L	518	Standard
	Sb	123	2208.8	28.0	**0.4480**	0.119	26.5	ug/L	127	Standard
	Ba	135	12894.8	1.5	**6.1480**	0.082	1.3	ug/L	45	Standard
	Ce	140	871.7	9.1				ug/L	298	Standard
>	**Tb**	159	974992.3	2.7				ug/L	977942	Standard
	Ho	165	33.3	48.2				ug/L	12	Standard
	Tl	203	98.0	7.7	**0.0064**	0.001	11.0	ug/L	31	Standard
	Tl	205	90.0	24.2	**0.0096**	0.003	27.9	ug/L	20	Standard
	Pb	206	392.7	2.7	**0.0066**	0.003	42.8	ug/L	286	Standard
	Pb	207	318.0	2.7	**0.0050**	0.001	23.7	ug/L	257	Standard
	Pb	208	1352.7	4.0	**0.0088**	0.002	18.5	ug/L	956	Standard
	U	238	56.0	29.8	**0.0046**	0.002	47.0	ug/L	17	Standard
>	**Bi**	209	576774.8	1.6				ug/L	557080	Standard

Sample ID: L1606084301SDL WG574570-02
Report Date/Time: Tuesday, June 28, 2016 20:19:17
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.6259**	18.396	173.1	mg/L	0	Standard
Mg	24	1335.1	0.6	**1.6748**	0.038	2.2	mg/L	32	Standard
K	39	83.3	27.1	**0.1320**	0.069	52.2	mg/L	30	Standard
Ca	43	66.7	4.3	**4.5227**	0.490	10.8	mg/L	48	Standard
Fe	54	158.9	15.8	**0.0030**	0.021	715.4	mg/L	159	Standard
Fe	57	226.7	20.8	**0.1225**	0.167	136.7	mg/L	195	Standard
> **Sc-1**	45	40586.0	2.0				mg/L	39543	Standard
Cl	35	41629.1	0.8				ug/L	27642	Standard
Kr	83	4.7	24.7				ug/L	2	Standard
Br	81	2016.8	5.8				ug/L	1647	Standard
P	31	19454.2	1.7				ug/L	15231	Standard
S	34	2286.8	4.6				ug/L	1845	Standard
Sr	88	108.3	9.6				ug/L	70	Standard
C	12	106.7	39.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	58.6	31.0				mg/L	12	Standard
Ho-1	165	33.3	48.2				mg/L	12	Standard
Er	166	30.0	57.7				mg/L	20	Standard
I	127	4460.7	6.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		103.486	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		103.697	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606084301SDL WG574570-02
Report Date/Time: Tuesday, June 28, 2016 20:19:17
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.200
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.535
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606084301SDL WG574570-02
Sample Date/Time: Tuesday, June 28, 2016 20:20:12
Number of Replicates: 3
Autosampler Position: 310
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84917.5	3.2				ug/L	81760	Standard
	Be	9	15.0	66.7	**0.0050**	0.010	200.3	ug/L	17	Standard
	Al	27	312447.8	1.9	**1.9378**	0.034	1.7	ug/L	1410	Standard
	Sc	45	41284.5	3.3				ug/L	39543	Standard
	Ti	47	45.7	48.8	**0.0117**	0.082	700.7	ug/L	39	Standard
	V	51	1463.2	5.6	**-0.0026**	0.007	263.9	ug/L	1415	Standard
	Cr	52	7659.3	1.8	**0.0098**	0.004	44.8	ug/L	6946	Standard
	Cr	53	423.3	16.3	**0.1274**	0.086	67.2	ug/L	328	Standard
	Mn	55	187375.9	0.8	**27.8813**	0.482	1.7	ug/L	823	Standard
	Co	59	2308.5	0.8	**0.2873**	0.008	2.8	ug/L	252	Standard
	Ni	60	1540.7	3.0	**0.6883**	0.022	3.2	ug/L	192	Standard
	Cu	65	216.7	7.3	**0.0197**	0.006	31.0	ug/L	141	Standard
	Zn	66	4319.6	1.9	**2.9936**	0.025	0.8	ug/L	491	Standard
>	**Ge**	72	531260.9	2.2				ug/L	515087	Standard
	As	75	-47.8	76.7	**-0.0014**	0.032	2343.8	ug/L	20	Standard
	Se	82	17.0	17.8	**0.0237**	0.023	96.9	ug/L	15	Standard
	Se-1	77	59.0	6.1	**0.0400**	0.030	74.0	ug/L	48	Standard
>	**Ga**	71	86.7	31.8				mg/L	90	Standard
	Rb	85	596.7	10.5				ug/L	23	Standard
	Y	89	453907.4	1.0				ug/L	448510	Standard
>	**Rh**	103	8.3	34.6				ug/L	15	Standard
	Mo	98	28.3	9.0	**-0.0090**	0.001	10.4	ug/L	11	Standard
	Ag	107	92.0	2.2	**-0.0030**	0.000	8.2	ug/L	134	Standard
	Cd	111	15.3	15.1	**0.0040**	0.001	27.2	mg/L	7	Standard
	Cd	114	38.2	13.4	**0.0022**	0.001	41.3	ug/L	16	Standard
>	**In**	115	632674.4	0.8				ug/L	611117	Standard
	Sn	118	625.0	9.7	**0.0002**	0.010	5977.1	ug/L	518	Standard
	Sb	123	614.9	31.7	**0.1128**	0.039	34.5	ug/L	127	Standard
	Ba	135	2594.2	0.9	**1.1887**	0.008	0.6	ug/L	45	Standard
	Ce	140	198.3	20.2				ug/L	298	Standard
>	**Tb**	159	988615.6	0.9				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	26.7	16.9	**-0.0018**	0.000	28.3	ug/L	31	Standard
	Tl	205	16.7	75.5	**-0.0001**	0.002	1448.6	ug/L	20	Standard
	Pb	206	345.0	3.1	**-0.0028**	0.002	58.3	ug/L	286	Standard
	Pb	207	303.3	7.6	**0.0012**	0.004	352.8	ug/L	257	Standard
	Pb	208	1152.7	2.7	**-0.0023**	0.002	71.0	ug/L	956	Standard
	U	238	16.0	21.7	**-0.0008**	0.000	54.8	ug/L	17	Standard
>	**Bi**	209	585233.9	0.6				ug/L	557080	Standard

Sample ID: L1606084301SDL WG574570-02
Report Date/Time: Tuesday, June 28, 2016 20:22:29
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	291.7	12.4	**0.3140**	0.035	11.1	mg/L	32	Standard
K	39	48.3	6.0	**0.0153**	0.013	86.9	mg/L	30	Standard
Ca	43	35.0	28.6	**-0.2811**	1.306	464.6	mg/L	48	Standard
Fe	54	153.8	34.6	**-0.0038**	0.045	1180.2	mg/L	159	Standard
Fe	57	170.0	8.8	**-0.0754**	0.036	48.4	mg/L	195	Standard
Sc-1	45	41284.5	3.3				mg/L	39543	Standard
Cl	35	41099.7	0.8				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	1820.1	5.7				ug/L	1647	Standard
P	31	19375.7	1.6				ug/L	15231	Standard
S	34	2545.2	4.1				ug/L	1845	Standard
Sr	88	58.3	27.6				ug/L	70	Standard
C	12	110.0	39.6				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	28.2	93.4				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	36.7	15.7				mg/L	20	Standard
I	127	1790.1	13.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.862	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.140	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606084301SDL WG574570-02
Report Date/Time: Tuesday, June 28, 2016 20:22:29
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.528
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.054
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 20:23:25
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	82271.2	2.6				ug/L	81760	Standard
	Be	9	51878.7	2.6	**51.9060**	0.474	0.9	ug/L	17	Standard
	Al	27	7912369.8	3.4	**50.9914**	0.481	0.9	ug/L	1410	Standard
	Sc	45	40149.8	1.3				ug/L	39543	Standard
	Ti	47	25445.3	0.9	**101.2486**	0.646	0.6	ug/L	39	Standard
	V	51	355257.7	1.2	**50.4501**	0.253	0.5	ug/L	1415	Standard
	Cr	52	364176.4	2.0	**50.7897**	0.304	0.6	ug/L	6946	Standard
	Cr	53	45239.1	0.4	**50.7395**	0.592	1.2	ug/L	328	Standard
	Mn	55	333791.7	0.4	**50.8845**	0.584	1.1	ug/L	823	Standard
	Co	59	367088.8	1.7	**51.7480**	0.425	0.8	ug/L	252	Standard
	Ni	60	95399.1	1.8	**50.0959**	0.391	0.8	ug/L	192	Standard
	Cu	65	97309.2	2.1	**50.6606**	0.365	0.7	ug/L	141	Standard
	Zn	66	56912.4	2.6	**50.6269**	0.602	1.2	ug/L	491	Standard
>	Ge	72	519736.7	1.5				ug/L	515087	Standard
	As	75	56554.6	1.5	**49.8917**	0.034	0.1	ug/L	20	Standard
	Se	82	5843.7	2.8	**50.1784**	0.757	1.5	ug/L	15	Standard
	Se-1	77	3965.5	1.4	**51.4490**	0.315	0.6	ug/L	48	Standard
>	Ga	71	118.3	13.6				mg/L	90	Standard
	Rb	85	848.4	6.2				ug/L	23	Standard
	Y	89	451425.6	1.6				ug/L	448510	Standard
>	Rh	103	31.7	71.2				ug/L	15	Standard
	Mo	98	272060.7	0.6	**99.5752**	0.551	0.6	ug/L	11	Standard
	Ag	107	339081.9	1.7	**49.9711**	1.119	2.2	ug/L	134	Standard
	Cd	111	106686.1	0.8	**49.3843**	0.588	1.2	mg/L	7	Standard
	Cd	114	259655.5	1.7	**49.4494**	1.121	2.3	ug/L	16	Standard
>	In	115	617617.8	0.6				ug/L	611117	Standard
	Sn	118	302897.0	2.2	**50.5905**	1.415	2.8	ug/L	518	Standard
	Sb	123	239436.4	1.3	**50.2037**	0.916	1.8	ug/L	127	Standard
	Ba	135	101930.1	0.6	**48.8299**	0.530	1.1	ug/L	45	Standard
	Ce	140	63.3	9.1				ug/L	298	Standard
>	Tb	159	978625.5	0.9				ug/L	977942	Standard
	Ho	165	8.3	69.3				ug/L	12	Standard
	Tl	203	434700.2	0.8	**50.4033**	0.455	0.9	ug/L	31	Standard
	Tl	205	381183.1	1.9	**50.8972**	0.874	1.7	ug/L	20	Standard
	Pb	206	277544.1	1.7	**50.1113**	0.892	1.8	ug/L	286	Standard
	Pb	207	250965.9	2.0	**50.2393**	1.021	2.0	ug/L	257	Standard
	Pb	208	994830.3	0.3	**51.4752**	0.269	0.5	ug/L	956	Standard
	U	238	372750.7	1.1	**50.8268**	0.592	1.2	ug/L	17	Standard
>	Bi	209	570174.9	0.2				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 20:25:42
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4207.3	2.1	**5.4598**	0.160	2.9	mg/L	32	Standard
K	39	1353.4	1.9	**4.3620**	0.034	0.8	mg/L	30	Standard
Ca	43	68.3	23.5	**4.8558**	2.284	47.0	mg/L	48	Standard
Fe	54	6096.4	4.5	**5.2968**	0.235	4.4	mg/L	159	Standard
Fe	57	1726.8	6.1	**5.1204**	0.425	8.3	mg/L	195	Standard
Sc-1	45	40149.8	1.3				mg/L	39543	Standard
Cl	35	38223.5	2.9				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	1710.1	6.5				ug/L	1647	Standard
P	31	20984.5	3.6				ug/L	15231	Standard
S	34	2836.9	3.0				ug/L	1845	Standard
Sr	88	73.3	32.2				ug/L	70	Standard
C	12	176.7	8.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	26.3	80.0				mg/L	12	Standard
Ho-1	165	8.3	69.3				mg/L	12	Standard
Er	166	6.7	86.6				mg/L	20	Standard
I	127	1388.4	24.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	103.812		
Al	27	101.983		
Sc	45			
Ti	47	101.249		
V	51	100.900		
Cr	52	101.579		
Cr	53			
Mn	55	101.769		
Co	59	103.496		
Ni	60	100.192		
Cu	65	101.321		
Zn	66	101.254		
Ge	72		100.903	
As	75	99.783		
Se	82	100.357		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 20:25:42
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	99.575		
Ag	107	99.942		
Cd	111	98.769		
Cd	114			
In	115			101.064
Sn	118	101.181		
Sb	123	100.407		
Ba	135	97.660		
Ce	140			
Tb	159			
Ho	165			
Tl	203	100.807		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.950		
U	238	101.654		
Bi	209			102.351
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 20:25:42
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 20:26:37
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77291.7	2.7				ug/L	81760	Standard
	Be	9	25.0	40.0	**0.0170**	0.011	63.1	ug/L	17	Standard
	Al	27	2576.9	24.9	**0.0032**	0.004	126.1	ug/L	1410	Standard
	Sc	45	39050.3	1.7				ug/L	39543	Standard
	Ti	47	33.7	13.4	**-0.0246**	0.019	78.0	ug/L	39	Standard
	V	51	1167.6	8.0	**-0.0323**	0.013	41.5	ug/L	1415	Standard
	Cr	52	6011.5	1.2	**-0.1610**	0.012	7.3	ug/L	6946	Standard
	Cr	53	398.3	5.1	**0.1293**	0.022	17.0	ug/L	328	Standard
	Mn	55	883.7	9.4	**-0.0071**	0.013	183.2	ug/L	823	Standard
	Co	59	243.0	13.8	**0.0045**	0.005	107.7	ug/L	252	Standard
	Ni	60	165.0	10.4	**-0.0141**	0.009	65.1	ug/L	192	Standard
	Cu	65	129.3	4.3	**-0.0202**	0.003	14.1	ug/L	141	Standard
	Zn	66	451.0	1.9	**-0.4007**	0.009	2.3	ug/L	491	Standard
>	**Ge**	72	496420.0	0.5				ug/L	515087	Standard
	As	75	-52.8	22.7	**-0.0085**	0.011	133.2	ug/L	20	Standard
	Se	82	7.8	31.0	**-0.0491**	0.022	44.9	ug/L	15	Standard
	Se-1	77	52.0	3.8	**-0.0027**	0.026	948.5	ug/L	48	Standard
>	**Ga**	71	61.7	18.7				mg/L	90	Standard
	Rb	85	18.3	103.3				ug/L	23	Standard
	Y	89	442161.2	1.2				ug/L	448510	Standard
>	**Rh**	103	6.7	43.3				ug/L	15	Standard
	Mo	98	281.8	35.5	**0.0839**	0.034	41.1	ug/L	11	Standard
	Ag	107	122.3	22.2	**0.0018**	0.004	200.6	ug/L	134	Standard
	Cd	111	14.5	44.1	**0.0038**	0.003	74.5	mg/L	7	Standard
	Cd	114	43.1	30.3	**0.0033**	0.002	70.6	ug/L	16	Standard
>	**In**	115	615841.7	2.0				ug/L	611117	Standard
	Sn	118	753.4	17.4	**0.0248**	0.024	98.6	ug/L	518	Standard
	Sb	123	596.8	25.3	**0.1131**	0.034	30.2	ug/L	127	Standard
	Ba	135	38.0	16.0	**-0.0069**	0.003	39.0	ug/L	45	Standard
	Ce	140	8.3	91.7				ug/L	298	Standard
>	**Tb**	159	956084.4	3.2				ug/L	977942	Standard
	Ho	165	5.0	100.0				ug/L	12	Standard
	Tl	203	73.0	30.8	**0.0039**	0.003	64.7	ug/L	31	Standard
	Tl	205	60.0	44.1	**0.0059**	0.003	58.6	ug/L	20	Standard
	Pb	206	307.3	5.4	**-0.0063**	0.002	32.6	ug/L	286	Standard
	Pb	207	254.3	1.8	**-0.0055**	0.000	8.7	ug/L	257	Standard
	Pb	208	1093.3	8.1	**-0.0021**	0.004	180.6	ug/L	956	Standard
	U	238	62.0	42.0	**0.0057**	0.004	62.4	ug/L	17	Standard
>	**Bi**	209	552726.9	1.8				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 20:28:53
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.8094**	18.714	173.1	mg/L	0	Standard
Mg	24	20.0	100.0	**-0.0300**	0.027	91.7	mg/L	32	Standard
K	39	31.7	50.8	**-0.0336**	0.053	157.9	mg/L	30	Standard
Ca	43	16.7	17.3	**-2.7773**	0.490	17.6	mg/L	48	Standard
Fe	54	136.0	26.0	**-0.0122**	0.034	282.8	mg/L	159	Standard
Fe	57	233.3	19.3	**0.1724**	0.147	85.5	mg/L	195	Standard
> **Sc-1**	45	39050.3	1.7				mg/L	39543	Standard
Cl	35	37106.7	0.9				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	1596.8	5.3				ug/L	1647	Standard
P	31	19597.7	1.6				ug/L	15231	Standard
S	34	2643.6	3.6				ug/L	1845	Standard
Sr	88	53.3	61.0				ug/L	70	Standard
C	12	110.0	32.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	0	Standard
Dy	164	9.4	104.3				mg/L	12	Standard
Ho-1	165	5.0	100.0				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	1173.4	25.6				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		96.376	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 20:28:53
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.773
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.219
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 20:28:53
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606079401
Sample Date/Time: Tuesday, June 28, 2016 20:29:50
Number of Replicates: 3
Autosampler Position: 311
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76244.9	4.1				ug/L	81760	Standard
	Be	9	31.7	63.8	**0.0239**	0.020	83.5	ug/L	17	Standard
	Al	27	4859544.3	1.6	**33.8379**	1.720	5.1	ug/L	1410	Standard
	Sc	45	41316.3	3.1				ug/L	39543	Standard
	Ti	47	1613.1	8.6	**6.3222**	0.521	8.2	ug/L	39	Standard
	V	51	5296.6	3.8	**0.5556**	0.022	4.0	ug/L	1415	Standard
	Cr	52	9948.6	1.0	**0.3727**	0.012	3.3	ug/L	6946	Standard
	Cr	53	1300.1	3.5	**1.1398**	0.032	2.8	ug/L	328	Standard
	Mn	55	10830112.5	0.4	**1671.2513**	23.447	1.4	ug/L	823	Standard
	Co	59	98900.4	0.9	**14.0508**	0.072	0.5	ug/L	252	Standard
	Ni	60	5918.2	0.8	**3.0389**	0.039	1.3	ug/L	192	Standard
	Cu	65	1492.7	4.1	**0.6949**	0.022	3.2	ug/L	141	Standard
	Zn	66	3606.1	2.5	**2.4634**	0.043	1.7	ug/L	491	Standard
>	**Ge**	72	514893.7	1.4				ug/L	515087	Standard
	As	75	2838.5	0.7	**2.5661**	0.029	1.1	ug/L	20	Standard
	Se	82	67.5	5.6	**0.4677**	0.041	8.7	ug/L	15	Standard
	Se-1	77	72.7	2.9	**0.2462**	0.031	12.7	ug/L	48	Standard
>	**Ga**	71	795.0	2.5				mg/L	90	Standard
	Rb	85	32504.8	2.3				ug/L	23	Standard
	Y	89	435925.3	2.8				ug/L	448510	Standard
>	**Rh**	103	93.3	3.1				ug/L	15	Standard
	Mo	98	343.6	1.5	**0.1114**	0.005	4.3	ug/L	11	Standard
	Ag	107	117.7	11.3	**0.0018**	0.002	118.8	ug/L	134	Standard
	Cd	111	40.7	19.8	**0.0166**	0.004	21.3	mg/L	7	Standard
	Cd	114	97.5	38.9	**0.0145**	0.008	54.8	ug/L	16	Standard
>	**In**	115	595657.1	2.2				ug/L	611117	Standard
	Sn	118	961.7	5.9	**0.0649**	0.008	12.2	ug/L	518	Standard
	Sb	123	345.9	22.0	**0.0622**	0.015	24.5	ug/L	127	Standard
	Ba	135	208994.2	0.7	**103.8626**	2.069	2.0	ug/L	45	Standard
	Ce	140	33943.0	0.8				ug/L	298	Standard
>	**Tb**	159	940708.8	1.0				ug/L	977942	Standard
	Ho	165	610.0	7.0				ug/L	12	Standard
	Tl	203	714.0	4.8	**0.0811**	0.003	4.2	ug/L	31	Standard
	Tl	205	651.7	8.6	**0.0881**	0.009	9.8	ug/L	20	Standard
	Pb	206	1354.7	2.2	**0.1905**	0.003	1.6	ug/L	286	Standard
	Pb	207	1119.4	1.4	**0.1747**	0.005	2.8	ug/L	257	Standard
	Pb	208	4748.3	2.1	**0.1947**	0.004	1.9	ug/L	956	Standard
	U	238	416.3	2.3	**0.0559**	0.002	3.0	ug/L	17	Standard
>	**Bi**	209	549596.9	1.2				ug/L	557080	Standard

Sample ID: L1606079401
Report Date/Time: Tuesday, June 28, 2016 20:32:07
Page 1

Approved: June 29, 2016

Na	23	6.7	114.6	**41.1239**	47.883	116.4	mg/L	0	Standard
Mg	24	3483.7	2.6	**4.3856**	0.231	5.3	mg/L	32	Standard
K	39	496.7	14.5	**1.4639**	0.222	15.1	mg/L	30	Standard
Ca	43	190.0	16.0	**22.2436**	4.083	18.4	mg/L	48	Standard
Fe	54	7377.0	1.5	**6.2559**	0.190	3.0	mg/L	159	Standard
Fe	57	2103.5	3.2	**6.1801**	0.414	6.7	mg/L	195	Standard
> **Sc-1**	45	41316.3	3.1				mg/L	39543	Standard
Cl	35	35106.6	2.5				ug/L	27642	Standard
Kr	83	2.3	107.9				ug/L	2	Standard
Br	81	13476.0	5.7				ug/L	1647	Standard
P	31	22598.5	5.8				ug/L	15231	Standard
S	34	2571.9	2.2				ug/L	1845	Standard
Sr	88	433.3	9.8				ug/L	70	Standard
C	12	150.0	35.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	882.4	15.5				mg/L	12	Standard
Ho-1	165	610.0	7.0				mg/L	12	Standard
Er	166	576.7	6.6				mg/L	20	Standard
I	127	50022.6	4.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.254	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.963	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606079401
Report Date/Time: Tuesday, June 28, 2016 20:32:07
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.470
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	98.657
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606079401
Report Date/Time: Tuesday, June 28, 2016 20:32:07
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606079402
Sample Date/Time: Tuesday, June 28, 2016 20:33:01
Number of Replicates: 3
Autosampler Position: 312
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75678.5	1.2				ug/L	81760	Standard
	Be	9	111.7	13.7	**0.1118**	0.017	15.0	ug/L	17	Standard
	Al	27	1429595.1	1.1	**10.0072**	0.228	2.3	ug/L	1410	Standard
	Sc	45	39138.8	1.6				ug/L	39543	Standard
	Ti	47	278.3	21.3	**0.9895**	0.212	21.4	ug/L	39	Standard
	V	51	1557.1	6.8	**0.0252**	0.009	34.9	ug/L	1415	Standard
	Cr	52	8352.6	6.5	**0.1848**	0.041	22.0	ug/L	6946	Standard
	Cr	53	1770.1	12.2	**1.7430**	0.191	11.0	ug/L	328	Standard
	Mn	55	34175.5	2.4	**5.3156**	0.224	4.2	ug/L	823	Standard
	Co	59	3266.0	5.1	**0.4499**	0.010	2.1	ug/L	252	Standard
	Ni	60	1953.8	3.9	**0.9695**	0.017	1.8	ug/L	192	Standard
	Cu	65	1358.7	2.9	**0.6499**	0.011	1.6	ug/L	141	Standard
	Zn	66	4870.4	2.4	**3.7758**	0.058	1.5	ug/L	491	Standard
>	**Ge**	72	497312.9	3.3				ug/L	515087	Standard
	As	75	54.7	94.2	**0.0915**	0.049	53.2	ug/L	20	Standard
	Se	82	40.7	27.3	**0.2488**	0.112	45.1	ug/L	15	Standard
	Se-1	77	77.0	9.4	**0.3389**	0.077	22.7	ug/L	48	Standard
>	**Ga**	71	155.0	28.1				mg/L	90	Standard
	Rb	85	30158.3	1.3				ug/L	23	Standard
	Y	89	432228.9	1.1				ug/L	448510	Standard
>	**Rh**	103	11.7	49.5				ug/L	15	Standard
	Mo	98	72.4	21.3	**0.0082**	0.006	70.5	ug/L	11	Standard
	Ag	107	118.3	24.0	**0.0017**	0.004	249.1	ug/L	134	Standard
	Cd	111	44.6	10.6	**0.0183**	0.002	12.1	mg/L	7	Standard
	Cd	114	123.7	31.0	**0.0193**	0.007	38.8	ug/L	16	Standard
>	**In**	115	600554.1	0.2				ug/L	611117	Standard
	Sn	118	733.4	8.7	**0.0243**	0.011	44.6	ug/L	518	Standard
	Sb	123	231.2	28.1	**0.0370**	0.014	37.8	ug/L	127	Standard
	Ba	135	65483.6	1.3	**32.2516**	0.425	1.3	ug/L	45	Standard
	Ce	140	2745.2	3.0				ug/L	298	Standard
>	**Tb**	159	936903.6	1.4				ug/L	977942	Standard
	Ho	165	75.0	17.6				ug/L	12	Standard
	Tl	203	221.3	13.5	**0.0222**	0.003	15.7	ug/L	31	Standard
	Tl	205	186.7	13.2	**0.0239**	0.003	14.3	ug/L	20	Standard
	Pb	206	523.0	7.4	**0.0358**	0.006	17.6	ug/L	286	Standard
	Pb	207	464.7	6.5	**0.0400**	0.008	19.0	ug/L	257	Standard
	Pb	208	1837.7	5.9	**0.0397**	0.005	13.6	ug/L	956	Standard
	U	238	125.7	23.1	**0.0151**	0.004	27.1	ug/L	17	Standard
>	**Bi**	209	541937.8	1.7				ug/L	557080	Standard

Sample ID: L1606079402
Report Date/Time: Tuesday, June 28, 2016 20:35:18
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	2726.9	0.4	**3.6111**	0.076	2.1	mg/L	32	Standard
K	39	506.7	5.1	**1.5885**	0.091	5.7	mg/L	30	Standard
Ca	43	36.7	15.7	**0.2847**	0.917	322.3	mg/L	48	Standard
Fe	54	182.6	18.3	**0.0302**	0.033	110.1	mg/L	159	Standard
Fe	57	203.3	25.6	**0.0706**	0.187	264.3	mg/L	195	Standard
> **Sc-1**	45	39138.8	1.6				mg/L	39543	Standard
Cl	35	34567.4	3.2				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	6558.1	13.2				ug/L	1647	Standard
P	31	21156.4	3.6				ug/L	15231	Standard
S	34	2326.8	9.7				ug/L	1845	Standard
Sr	88	76.7	13.6				ug/L	70	Standard
C	12	236.7	44.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	150.6	9.6				mg/L	12	Standard
Ho-1	165	75.0	17.6				mg/L	12	Standard
Er	166	56.7	40.8				mg/L	20	Standard
I	127	9424.6	4.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		92.562	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.549	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606079402
Report Date/Time: Tuesday, June 28, 2016 20:35:18
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.272
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.282
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606079403
Sample Date/Time: Tuesday, June 28, 2016 20:36:13
Number of Replicates: 3
Autosampler Position: 313
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76753.9	0.3				ug/L	81760	Standard
	Be	9	53.3	32.9	**0.0475**	0.019	39.7	ug/L	17	Standard
	Al	27	2019995.9	2.1	**13.9455**	0.321	2.3	ug/L	1410	Standard
	Sc	45	39648.5	2.7				ug/L	39543	Standard
	Ti	47	884.4	21.1	**3.3748**	0.755	22.4	ug/L	39	Standard
	V	51	4216.0	3.5	**0.3970**	0.022	5.5	ug/L	1415	Standard
	Cr	52	11350.6	1.0	**0.5656**	0.020	3.6	ug/L	6946	Standard
	Cr	53	3082.0	1.2	**3.1525**	0.043	1.4	ug/L	328	Standard
	Mn	55	1462715.3	1.2	**224.3814**	2.702	1.2	ug/L	823	Standard
	Co	59	3343.0	3.0	**0.4421**	0.013	2.9	ug/L	252	Standard
	Ni	60	3343.7	2.8	**1.6618**	0.047	2.8	ug/L	192	Standard
	Cu	65	1050.7	3.0	**0.4594**	0.015	3.2	ug/L	141	Standard
	Zn	66	4988.5	1.4	**3.7017**	0.077	2.1	ug/L	491	Standard
>	**Ge**	72	517603.4	0.3				ug/L	515087	Standard
	As	75	193.9	14.0	**0.2119**	0.024	11.1	ug/L	20	Standard
	Se	82	57.1	5.5	**0.3743**	0.026	6.9	ug/L	15	Standard
	Se-1	77	124.7	7.3	**0.9281**	0.125	13.4	ug/L	48	Standard
>	**Ga**	71	286.7	21.7				mg/L	90	Standard
	Rb	85	31280.6	0.6				ug/L	23	Standard
	Y	89	445698.9	0.7				ug/L	448510	Standard
>	**Rh**	103	16.7	45.8				ug/L	15	Standard
	Mo	98	229.5	3.6	**0.0662**	0.004	6.0	ug/L	11	Standard
	Ag	107	129.3	14.3	**0.0031**	0.003	91.8	ug/L	134	Standard
	Cd	111	79.6	15.3	**0.0346**	0.006	17.8	mg/L	7	Standard
	Cd	114	219.2	32.9	**0.0375**	0.014	38.3	ug/L	16	Standard
>	**In**	115	608521.7	1.3				ug/L	611117	Standard
	Sn	118	713.4	5.3	**0.0192**	0.007	34.1	ug/L	518	Standard
	Sb	123	215.5	21.1	**0.0330**	0.009	27.4	ug/L	127	Standard
	Ba	135	73971.5	1.1	**35.9586**	0.191	0.5	ug/L	45	Standard
	Ce	140	14728.8	3.9				ug/L	298	Standard
>	**Tb**	159	956871.9	1.4				ug/L	977942	Standard
	Ho	165	336.7	6.7				ug/L	12	Standard
	Tl	203	279.3	8.0	**0.0283**	0.003	10.3	ug/L	31	Standard
	Tl	205	295.0	11.9	**0.0379**	0.004	11.8	ug/L	20	Standard
	Pb	206	724.7	2.0	**0.0702**	0.002	2.4	ug/L	286	Standard
	Pb	207	605.7	5.7	**0.0660**	0.008	12.2	ug/L	257	Standard
	Pb	208	2584.1	3.2	**0.0763**	0.006	7.3	ug/L	956	Standard
	U	238	605.0	8.8	**0.0814**	0.008	10.1	ug/L	17	Standard
>	**Bi**	209	558048.4	1.1				ug/L	557080	Standard

Sample ID: L1606079403
Report Date/Time: Tuesday, June 28, 2016 20:38:30
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	4734.1	2.7	**6.2297**	0.190	3.1	mg/L	32	Standard
K	39	378.3	13.8	**1.1326**	0.165	14.5	mg/L	30	Standard
Ca	43	55.0	18.2	**3.0097**	1.672	55.5	mg/L	48	Standard
Fe	54	512.5	3.2	**0.3260**	0.027	8.3	mg/L	159	Standard
Fe	57	288.3	19.9	**0.3473**	0.203	58.5	mg/L	195	Standard
> **Sc-1**	45	39648.5	2.7				mg/L	39543	Standard
Cl	35	35720.7	1.2				ug/L	27642	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	10650.4	4.7				ug/L	1647	Standard
P	31	21555.3	1.7				ug/L	15231	Standard
S	34	2313.5	4.4				ug/L	1845	Standard
Sr	88	90.0	14.7				ug/L	70	Standard
C	12	206.7	17.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0					mg/L	0	Standard
Dy	164	611.8	5.5				mg/L	12	Standard
Ho-1	165	336.7	6.7				mg/L	12	Standard
Er	166	310.0	11.6				mg/L	20	Standard
I	127	51010.9	6.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.877	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.489	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606079403
Report Date/Time: Tuesday, June 28, 2016 20:38:30
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	99.575	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	100.174	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606079403
Report Date/Time: Tuesday, June 28, 2016 20:38:30
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606079404
Sample Date/Time: Tuesday, June 28, 2016 20:39:24
Number of Replicates: 3
Autosampler Position: 314
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77373.8	1.6				ug/L	81760	Standard
	Be	9	33.3	45.8	**0.0257**	0.016	61.7	ug/L	17	Standard
	Al	27	6530294.9	1.7	**44.7685**	1.484	3.3	ug/L	1410	Standard
	Sc	45	40462.3	2.9				ug/L	39543	Standard
	Ti	47	1753.8	22.2	**6.9932**	1.564	22.4	ug/L	39	Standard
	V	51	4913.4	4.4	**0.5113**	0.026	5.2	ug/L	1415	Standard
	Cr	52	10044.0	5.5	**0.4080**	0.064	15.7	ug/L	6946	Standard
	Cr	53	2381.9	7.8	**2.4159**	0.236	9.8	ug/L	328	Standard
	Mn	55	8647420.3	0.2	**1354.8198**	14.536	1.1	ug/L	823	Standard
	Co	59	169235.6	0.5	**24.4356**	0.161	0.7	ug/L	252	Standard
	Ni	60	12361.0	0.9	**6.5619**	0.023	0.4	ug/L	192	Standard
	Cu	65	806.0	0.9	**0.3401**	0.009	2.5	ug/L	141	Standard
	Zn	66	7577.2	1.7	**6.1942**	0.082	1.3	ug/L	491	Standard
>	**Ge**	72	507111.0	1.2				ug/L	515087	Standard
	As	75	265.8	16.5	**0.2802**	0.037	13.2	ug/L	20	Standard
	Se	82	64.6	10.0	**0.4504**	0.053	11.9	ug/L	15	Standard
	Se-1	77	107.0	10.8	**0.7228**	0.140	19.4	ug/L	48	Standard
>	**Ga**	71	598.3	9.8				mg/L	90	Standard
	Rb	85	32554.9	1.8				ug/L	23	Standard
	Y	89	447221.5	3.0				ug/L	448510	Standard
>	**Rh**	103	20.0	50.0				ug/L	15	Standard
	Mo	98	99.0	16.4	**0.0174**	0.006	35.0	ug/L	11	Standard
	Ag	107	155.7	70.8	**0.0069**	0.016	237.9	ug/L	134	Standard
	Cd	111	107.8	50.7	**0.0474**	0.026	54.2	mg/L	7	Standard
	Cd	114	265.0	52.4	**0.0460**	0.027	58.4	ug/L	16	Standard
>	**In**	115	613219.8	0.3				ug/L	611117	Standard
	Sn	118	870.0	6.6	**0.0447**	0.010	22.4	ug/L	518	Standard
	Sb	123	214.6	24.5	**0.0325**	0.011	34.6	ug/L	127	Standard
	Ba	135	118463.0	0.8	**57.1584**	0.254	0.4	ug/L	45	Standard
	Ce	140	43559.2	3.6				ug/L	298	Standard
>	**Tb**	159	953430.3	1.1				ug/L	977942	Standard
	Ho	165	631.7	5.3				ug/L	12	Standard
	Tl	203	508.0	24.8	**0.0560**	0.015	27.1	ug/L	31	Standard
	Tl	205	431.7	17.4	**0.0571**	0.010	17.4	ug/L	20	Standard
	Pb	206	1022.0	12.5	**0.1271**	0.024	19.0	ug/L	286	Standard
	Pb	207	848.0	9.6	**0.1173**	0.017	14.2	ug/L	257	Standard
	Pb	208	3317.1	5.1	**0.1168**	0.009	7.7	ug/L	956	Standard
	U	238	210.3	17.6	**0.0266**	0.005	19.3	ug/L	17	Standard
>	**Bi**	209	552632.4	0.7				ug/L	557080	Standard

Sample ID: L1606079404
Report Date/Time: Tuesday, June 28, 2016 20:41:41
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**20.7118**	17.947	86.6	mg/L	0	Standard
Mg	24	5764.4	1.5	**7.4467**	0.270	3.6	mg/L	32	Standard
K	39	563.3	10.2	**1.7170**	0.152	8.8	mg/L	30	Standard
Ca	43	75.0	17.6	**5.8105**	2.161	37.2	mg/L	48	Standard
Fe	54	999.2	5.7	**0.7481**	0.074	9.9	mg/L	159	Standard
Fe	57	490.0	3.5	**0.9936**	0.094	9.5	mg/L	195	Standard
Sc-1	45	40462.3	2.9				mg/L	39543	Standard
Cl	35	35374.6	1.6				ug/L	27642	Standard
Kr	83	2.7	78.1				ug/L	2	Standard
Br	81	12395.1	8.6				ug/L	1647	Standard
P	31	20597.3	1.4				ug/L	15231	Standard
S	34	2416.9	2.7				ug/L	1845	Standard
Sr	88	128.3	19.6				ug/L	70	Standard
C	12	236.7	2.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	1138.1	44.8				mg/L	12	Standard
Ho-1	165	631.7	5.3				mg/L	12	Standard
Er	166	596.7	6.3				mg/L	20	Standard
I	127	203787.2	10.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.635	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.452	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606079404
Report Date/Time: Tuesday, June 28, 2016 20:41:41
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.344
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.202
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606082202
Sample Date/Time: Tuesday, June 28, 2016 20:42:36
Number of Replicates: 3
Autosampler Position: 315
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79831.5	1.5				ug/L	81760	Standard
	Be	9	18.3	63.0	**0.0091**	0.012	129.5	ug/L	17	Standard
	Al	27	3326140.8	2.2	**22.0876**	0.529	2.4	ug/L	1410	Standard
	Sc	45	40903.5	2.4				ug/L	39543	Standard
	Ti	47	258.3	7.7	**0.8950**	0.111	12.4	ug/L	39	Standard
	V	51	1427.8	26.8	**0.0038**	0.061	1620.7	ug/L	1415	Standard
	Cr	52	8606.8	1.2	**0.2025**	0.023	11.3	ug/L	6946	Standard
	Cr	53	7031.6	3.4	**7.8173**	0.215	2.7	ug/L	328	Standard
	Mn	55	245282.9	2.4	**38.3872**	0.371	1.0	ug/L	823	Standard
	Co	59	360.0	5.7	**0.0208**	0.001	6.7	ug/L	252	Standard
	Ni	60	675.7	7.1	**0.2607**	0.030	11.7	ug/L	192	Standard
	Cu	65	319.3	4.0	**0.0805**	0.011	13.1	ug/L	141	Standard
	Zn	66	1777.8	5.5	**0.8236**	0.060	7.3	ug/L	491	Standard
>	**Ge**	72	505790.0	3.0				ug/L	515087	Standard
	As	75	361.2	27.3	**0.3676**	0.091	24.7	ug/L	20	Standard
	Se	82	151.2	10.8	**1.2184**	0.141	11.6	ug/L	15	Standard
	Se-1	77	286.7	5.2	**3.1548**	0.107	3.4	ug/L	48	Standard
>	**Ga**	71	148.3	15.6				mg/L	90	Standard
	Rb	85	12600.2	4.8				ug/L	23	Standard
	Y	89	440998.1	2.4				ug/L	448510	Standard
>	**Rh**	103	48.3	6.0				ug/L	15	Standard
	Mo	98	329.5	7.2	**0.1031**	0.007	6.5	ug/L	11	Standard
	Ag	107	82.7	3.0	**-0.0039**	0.000	4.5	ug/L	134	Standard
	Cd	111	39.1	14.9	**0.0154**	0.002	16.0	mg/L	7	Standard
	Cd	114	93.5	42.9	**0.0131**	0.008	57.9	ug/L	16	Standard
>	**In**	115	609428.7	1.7				ug/L	611117	Standard
	Sn	118	616.7	6.1	**0.0027**	0.008	296.4	ug/L	518	Standard
	Sb	123	172.8	22.7	**0.0238**	0.008	32.2	ug/L	127	Standard
	Ba	135	142178.1	1.8	**69.0335**	0.137	0.2	ug/L	45	Standard
	Ce	140	378.3	16.5				ug/L	298	Standard
>	**Tb**	159	949587.4	3.3				ug/L	977942	Standard
	Ho	165	25.0	20.0				ug/L	12	Standard
	Tl	203	147.7	1.7	**0.0134**	0.001	4.1	ug/L	31	Standard
	Tl	205	166.7	29.1	**0.0214**	0.007	33.5	ug/L	20	Standard
	Pb	206	309.7	2.3	**-0.0043**	0.001	31.8	ug/L	286	Standard
	Pb	207	262.7	4.1	**-0.0022**	0.002	83.4	ug/L	257	Standard
	Pb	208	1080.7	3.0	**-0.0011**	0.001	98.3	ug/L	956	Standard
	U	238	30.0	28.9	**0.0014**	0.001	94.3	ug/L	17	Standard
>	**Bi**	209	537934.0	1.4				ug/L	557080	Standard

Sample ID: L1606082202
Report Date/Time: Tuesday, June 28, 2016 20:44:53
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.6139**	18.375	173.1	mg/L	0	Standard
Mg	24	32371.2	3.9	**41.6322**	2.388	5.7	mg/L	32	Standard
K	39	260.0	20.1	**0.7087**	0.176	24.8	mg/L	30	Standard
Ca	43	58.3	19.8	**3.1978**	1.533	48.0	mg/L	48	Standard
Fe	54	1294.3	5.1	**0.9965**	0.083	8.3	mg/L	159	Standard
Fe	57	445.0	8.5	**0.8270**	0.105	12.6	mg/L	195	Standard
Sc-1	45	40903.5	2.4				mg/L	39543	Standard
Cl	35	36431.1	1.6				ug/L	27642	Standard
Kr	83	0.3	173.2				ug/L	2	Standard
Br	81	35805.7	4.9				ug/L	1647	Standard
P	31	25079.1	3.1				ug/L	15231	Standard
S	34	2408.5	3.8				ug/L	1845	Standard
Sr	88	206.7	19.6				ug/L	70	Standard
C	12	176.7	45.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	28.1	3.0				mg/L	12	Standard
Ho-1	165	25.0	20.0				mg/L	12	Standard
Er	166	40.0	43.3				mg/L	20	Standard
I	127	201681.5	7.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		97.641	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		98.195	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606082202
Report Date/Time: Tuesday, June 28, 2016 20:44:53
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.724
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.563
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606082203
Sample Date/Time: Tuesday, June 28, 2016 20:45:47
Number of Replicates: 3
Autosampler Position: 316
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90191.0	1.0				ug/L	81760	Standard
	Be	9	23.3	68.9	**0.0116**	0.015	127.2	ug/L	17	Standard
	Al	27	11365406.0	0.9	**66.8332**	1.185	1.8	ug/L	1410	Standard
	Sc	45	43861.8	3.4				ug/L	39543	Standard
	Ti	47	293.7	8.7	**0.9575**	0.101	10.6	ug/L	39	Standard
	V	51	2061.8	11.0	**0.0753**	0.031	41.0	ug/L	1415	Standard
	Cr	52	35560.4	1.1	**3.7982**	0.065	1.7	ug/L	6946	Standard
	Cr	53	13979.8	6.4	**14.7966**	1.011	6.8	ug/L	328	Standard
	Mn	55	581441.7	2.8	**85.0974**	2.668	3.1	ug/L	823	Standard
	Co	59	11104.7	2.7	**1.4708**	0.043	3.0	ug/L	252	Standard
	Ni	60	4756.1	1.3	**2.2951**	0.027	1.2	ug/L	192	Standard
	Cu	65	585.0	3.7	**0.2018**	0.012	5.9	ug/L	141	Standard
	Zn	66	4219.9	2.2	**2.8317**	0.093	3.3	ug/L	491	Standard
>	**Ge**	72	541965.4	0.4				ug/L	515087	Standard
	As	75	1879.3	4.8	**1.6291**	0.081	5.0	ug/L	20	Standard
	Se	82	403.3	6.2	**3.2102**	0.210	6.5	ug/L	15	Standard
	Se-1	77	499.0	5.4	**5.5758**	0.322	5.8	ug/L	48	Standard
>	**Ga**	71	170.0	12.8				mg/L	90	Standard
	Rb	85	898318.0	1.7				ug/L	23	Standard
	Y	89	454033.5	1.6				ug/L	448510	Standard
>	**Rh**	103	158.3	15.6				ug/L	15	Standard
	Mo	98	390.2	3.3	**0.1238**	0.006	4.6	ug/L	11	Standard
	Ag	107	91.7	18.8	**-0.0027**	0.003	94.7	ug/L	134	Standard
	Cd	111	57.3	18.5	**0.0236**	0.005	19.9	mg/L	7	Standard
	Cd	114	162.3	23.5	**0.0261**	0.008	29.2	ug/L	16	Standard
>	**In**	115	617377.1	1.4				ug/L	611117	Standard
	Sn	118	971.7	10.6	**0.0609**	0.019	30.8	ug/L	518	Standard
	Sb	123	275.2	12.0	**0.0450**	0.008	16.8	ug/L	127	Standard
	Ba	135	754414.8	1.4	**361.7131**	4.988	1.4	ug/L	45	Standard
	Ce	140	4038.9	3.4				ug/L	298	Standard
>	**Tb**	159	972701.2	1.2				ug/L	977942	Standard
	Ho	165	60.0					ug/L	12	Standard
	Tl	203	156.0	13.2	**0.0138**	0.002	16.1	ug/L	31	Standard
	Tl	205	116.7	13.1	**0.0137**	0.002	15.9	ug/L	20	Standard
	Pb	206	526.0	2.4	**0.0341**	0.004	10.3	ug/L	286	Standard
	Pb	207	425.7	3.6	**0.0296**	0.004	13.9	ug/L	257	Standard
	Pb	208	1793.4	1.5	**0.0349**	0.002	5.6	ug/L	956	Standard
	U	238	113.7	2.7	**0.0130**	0.001	4.0	ug/L	17	Standard
>	**Bi**	209	555248.6	1.3				ug/L	557080	Standard

Sample ID: L1606082203
Report Date/Time: Tuesday, June 28, 2016 20:48:04
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**9.3851**	16.247	173.1	mg/L	0	Standard
Mg	24	35620.2	1.5	**42.7261**	1.684	3.9	mg/L	32	Standard
K	39	6901.6	2.1	**20.9070**	1.156	5.5	mg/L	30	Standard
Ca	43	215.0	12.3	**24.1234**	3.813	15.8	mg/L	48	Standard
Fe	54	3077.7	3.9	**2.3743**	0.067	2.8	mg/L	159	Standard
Fe	57	1036.7	11.9	**2.5246**	0.275	10.9	mg/L	195	Standard
> **Sc-1**	45	43861.8	3.4				mg/L	39543	Standard
Cl	35	37418.9	4.6				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	90214.6	3.4				ug/L	1647	Standard
P	31	25815.3	2.7				ug/L	15231	Standard
S	34	2615.2	3.8				ug/L	1845	Standard
Sr	88	703.3	5.7				ug/L	70	Standard
C	12	206.7	5.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	96.0	27.7				mg/L	12	Standard
Ho-1	165	60.0					mg/L	12	Standard
Er	166	83.3	6.9				mg/L	20	Standard
I	127	135411.8	5.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.312	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		105.218	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082203
Report Date/Time: Tuesday, June 28, 2016 20:48:04
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	101.024	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	99.671	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Ba 135 Upper, S, EEE	Ba	135	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606082302
Sample Date/Time: Tuesday, June 28, 2016 20:48:59
Number of Replicates: 3
Autosampler Position: 317
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	108277.4	2.9				ug/L	81760	Standard
	Be	9	45.0	44.4	**0.0244**	0.015	61.4	ug/L	17	Standard
	Al	27	155300534.6	2.6	**761.2127**	30.613	4.0	ug/L	1410	Standard
	Sc	45	45277.6	3.5				ug/L	39543	Standard
	Ti	47	403.3	5.8	**1.4273**	0.064	4.5	ug/L	39	Standard
	V	51	3112.6	16.0	**0.2334**	0.070	30.2	ug/L	1415	Standard
	Cr	52	14043.2	7.2	**0.9232**	0.105	11.4	ug/L	6946	Standard
	Cr	53	48930.1	15.5	**54.3111**	7.500	13.8	ug/L	328	Standard
	Mn	55	2748271.5	3.9	**416.1426**	7.847	1.9	ug/L	823	Standard
	Co	59	187853.6	1.8	**26.2312**	0.060	0.2	ug/L	252	Standard
	Ni	60	24191.7	7.1	**12.5042**	0.665	5.3	ug/L	192	Standard
	Cu	65	2608.9	7.7	**1.2568**	0.078	6.2	ug/L	141	Standard
	Zn	66	8434.7	3.5	**6.7297**	0.123	1.8	ug/L	491	Standard
>	**Ge**	72	524401.5	2.0				ug/L	515087	Standard
	As	75	6667.1	7.6	**5.8608**	0.340	5.8	ug/L	20	Standard
	Se	82	2463.0	3.4	**20.8902**	0.307	1.5	ug/L	15	Standard
	Se-1	77	5803.1	9.1	**74.8730**	5.475	7.3	ug/L	48	Standard
>	**Ga**	71	283.3	11.5				mg/L	90	Standard
	Rb	85	14086.5	1.8				ug/L	23	Standard
	Y	89	450377.0	4.3				ug/L	448510	Standard
>	**Rh**	103	1153.4	2.0				ug/L	15	Standard
	Mo	98	228.3	2.1	**0.0678**	0.002	2.3	ug/L	11	Standard
	Ag	107	100.7	7.3	**-0.0008**	0.001	114.3	ug/L	134	Standard
	Cd	111	85.9	11.5	**0.0385**	0.005	14.0	mg/L	7	Standard
	Cd	114	207.5	20.1	**0.0360**	0.007	20.1	ug/L	16	Standard
>	**In**	115	594734.3	2.8				ug/L	611117	Standard
	Sn	118	838.4	7.1	**0.0436**	0.007	15.0	ug/L	518	Standard
	Sb	123	341.7	13.9	**0.0614**	0.008	13.7	ug/L	127	Standard
	Ba	135	7531184.6	2.9	**3748.3398**	4.365	0.1	ug/L	45	Standard
	Ce	140	2656.9	7.0				ug/L	298	Standard
>	**Tb**	159	994456.1	1.2				ug/L	977942	Standard
	Ho	165	153.3	5.0				ug/L	12	Standard
	Tl	203	149.0	11.0	**0.0143**	0.002	16.0	ug/L	31	Standard
	Tl	205	131.7	44.0	**0.0171**	0.008	49.5	ug/L	20	Standard
	Pb	206	558.3	2.7	**0.0476**	0.002	3.6	ug/L	286	Standard
	Pb	207	480.7	4.4	**0.0481**	0.004	8.4	ug/L	257	Standard
	Pb	208	1800.4	1.7	**0.0423**	0.001	2.5	ug/L	956	Standard
	U	238	2848.6	3.5	**0.4250**	0.011	2.7	ug/L	17	Standard
>	**Bi**	209	517504.0	1.7				ug/L	557080	Standard

Sample ID: L1606082302
Report Date/Time: Tuesday, June 28, 2016 20:51:16
Page 1

Approved: June 29, 2016

Na	23	15.0	33.3	**83.8040**	26.058	31.1	mg/L	0	Standard
Mg	24	304856.8	3.2	**354.4066**	5.599	1.6	mg/L	32	Standard
K	39	523.3	14.8	**1.3989**	0.174	12.4	mg/L	30	Standard
Ca	43	1175.0	5.2	**150.6324**	9.407	6.2	mg/L	48	Standard
Fe	54	2706.9	4.6	**2.0025**	0.071	3.5	mg/L	159	Standard
Fe	57	1593.4	6.7	**4.0698**	0.154	3.8	mg/L	195	Standard
Sc-1	45	45277.6	3.5				mg/L	39543	Standard
Cl	35	40644.4	1.3				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	610984.3	3.4				ug/L	1647	Standard
P	31	30180.0	2.7				ug/L	15231	Standard
S	34	2781.9	5.2				ug/L	1845	Standard
Sr	88	4374.0	2.8				ug/L	70	Standard
C	12	443.3	22.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	166.1	26.6				mg/L	12	Standard
Ho-1	165	153.3	5.0				mg/L	12	Standard
Er	166	220.0	28.4				mg/L	20	Standard
I	127	589063.3	13.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.433	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.808	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082302
Report Date/Time: Tuesday, June 28, 2016 20:51:16
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.319
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	92.896
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606082302
Report Date/Time: Tuesday, June 28, 2016 20:51:16
Page 3

Approved: June 29, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606082303
Sample Date/Time: Tuesday, June 28, 2016 20:52:10
Number of Replicates: 3
Autosampler Position: 318
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93244.4	1.1				ug/L	81760	Standard
	Be	9	28.3	10.2	**0.0153**	0.003	16.9	ug/L	17	Standard
	Al	27	94744933.1	0.6	**538.9798**	6.462	1.2	ug/L	1410	Standard
	Sc	45	47113.2	4.5				ug/L	39543	Standard
	Ti	47	553.7	11.7	**2.0175**	0.206	10.2	ug/L	39	Standard
	V	51	4328.2	13.5	**0.4057**	0.092	22.6	ug/L	1415	Standard
	Cr	52	14490.9	2.4	**0.9864**	0.018	1.8	ug/L	6946	Standard
	Cr	53	43709.7	3.6	**48.5418**	1.565	3.2	ug/L	328	Standard
	Mn	55	8319616.3	2.0	**1259.6840**	13.722	1.1	ug/L	823	Standard
	Co	59	19767.9	2.2	**2.7309**	0.056	2.1	ug/L	252	Standard
	Ni	60	10810.2	4.4	**5.5279**	0.127	2.3	ug/L	192	Standard
	Cu	65	2456.2	2.4	**1.1782**	0.021	1.8	ug/L	141	Standard
	Zn	66	5426.0	2.4	**4.0320**	0.062	1.5	ug/L	491	Standard
>	**Ge**	72	524719.7	2.2				ug/L	515087	Standard
	As	75	3752.6	7.8	**3.3171**	0.255	7.7	ug/L	20	Standard
	Se	82	1738.6	3.2	**14.7077**	0.502	3.4	ug/L	15	Standard
	Se-1	77	4761.1	4.1	**61.3297**	2.628	4.3	ug/L	48	Standard
>	**Ga**	71	408.3	8.7				mg/L	90	Standard
	Rb	85	9714.8	3.7				ug/L	23	Standard
	Y	89	460620.0	3.3				ug/L	448510	Standard
>	**Rh**	103	891.7	10.4				ug/L	15	Standard
	Mo	98	389.5	4.4	**0.1260**	0.003	2.4	ug/L	11	Standard
	Ag	107	110.7	7.4	**0.0004**	0.001	272.1	ug/L	134	Standard
	Cd	111	332.6	3.0	**0.1538**	0.007	4.3	mg/L	7	Standard
	Cd	114	853.2	3.8	**0.1605**	0.008	5.1	ug/L	16	Standard
>	**In**	115	607039.3	2.6				ug/L	611117	Standard
	Sn	118	751.7	13.7	**0.0260**	0.016	61.8	ug/L	518	Standard
	Sb	123	376.4	10.5	**0.0674**	0.007	10.4	ug/L	127	Standard
	Ba	135	4357695.0	1.8	**2125.2206**	17.296	0.8	ug/L	45	Standard
	Ce	140	11329.2	2.7				ug/L	298	Standard
>	**Tb**	159	992575.1	2.0				ug/L	977942	Standard
	Ho	165	278.3	9.0				ug/L	12	Standard
	Tl	203	135.0	5.8	**0.0123**	0.001	6.8	ug/L	31	Standard
	Tl	205	110.0	27.6	**0.0138**	0.004	31.4	ug/L	20	Standard
	Pb	206	907.4	3.0	**0.1159**	0.005	4.7	ug/L	286	Standard
	Pb	207	750.0	6.5	**0.1064**	0.009	8.4	ug/L	257	Standard
	Pb	208	3103.1	4.1	**0.1154**	0.005	4.4	ug/L	956	Standard
	U	238	44356.2	2.2	**6.6145**	0.107	1.6	ug/L	17	Standard
>	**Bi**	209	521151.7	1.4				ug/L	557080	Standard

Sample ID: L1606082303
Report Date/Time: Tuesday, June 28, 2016 20:54:27
Page 1

Approved: June 29, 2016

Na	23	10.0	86.6	**54.6526**	47.439	86.8	mg/L	0	Standard
Mg	24	287663.3	3.0	**321.5385**	8.440	2.6	mg/L	32	Standard
K	39	226.7	14.2	**0.5032**	0.104	20.6	mg/L	30	Standard
Ca	43	1070.0	2.9	**131.3674**	10.299	7.8	mg/L	48	Standard
Fe	54	476.6	5.9	**0.2261**	0.038	16.9	mg/L	159	Standard
Fe	57	1008.4	3.2	**2.2367**	0.165	7.4	mg/L	195	Standard
> **Sc-1**	45	47113.2	4.5				mg/L	39543	Standard
Cl	35	40790.8	0.7				ug/L	27642	Standard
Kr	83	0.7	173.2				ug/L	2	Standard
Br	81	436089.5	3.9				ug/L	1647	Standard
P	31	27271.2	4.0				ug/L	15231	Standard
S	34	2836.9	4.2				ug/L	1845	Standard
Sr	88	2816.9	2.8				ug/L	70	Standard
C	12	376.7	12.5				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	20.0	86.6				mg/L	0	Standard
Dy	164	345.0	17.3				mg/L	12	Standard
Ho-1	165	278.3	9.0				mg/L	12	Standard
Er	166	243.3	18.5				mg/L	20	Standard
I	127	900084.0	10.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		114.046	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.870	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082303
Report Date/Time: Tuesday, June 28, 2016 20:54:27
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.333
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		93.551
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606082303
Report Date/Time: Tuesday, June 28, 2016 20:54:27
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606082304
Sample Date/Time: Tuesday, June 28, 2016 20:55:22
Number of Replicates: 3
Autosampler Position: 319
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74476.1	3.3				ug/L	81760	Standard
	Be	9	15.0	115.5	**0.0072**	0.020	276.2	ug/L	17	Standard
	Al	27	20080706.8	2.9	**143.0136**	1.080	0.8	ug/L	1410	Standard
	Sc	45	42894.0	3.7				ug/L	39543	Standard
	Ti	47	325.7	10.6	**1.1562**	0.137	11.8	ug/L	39	Standard
	V	51	1429.9	29.9	**0.0011**	0.063	5474.1	ug/L	1415	Standard
	Cr	52	7720.6	1.3	**0.0621**	0.011	18.4	ug/L	6946	Standard
	Cr	53	9723.1	6.1	**10.8363**	0.676	6.2	ug/L	328	Standard
	Mn	55	4663144.1	0.8	**725.7255**	4.778	0.7	ug/L	823	Standard
	Co	59	1828.4	1.1	**0.2313**	0.003	1.3	ug/L	252	Standard
	Ni	60	1442.4	3.3	**0.6679**	0.026	3.9	ug/L	192	Standard
	Cu	65	719.0	5.1	**0.2910**	0.019	6.5	ug/L	141	Standard
	Zn	66	7525.5	1.3	**6.1013**	0.101	1.7	ug/L	491	Standard
>	**Ge**	72	510419.4	0.4				ug/L	515087	Standard
	As	75	3004.5	4.7	**2.7368**	0.118	4.3	ug/L	20	Standard
	Se	82	527.6	3.3	**4.5049**	0.149	3.3	ug/L	15	Standard
	Se-1	77	877.0	5.5	**11.0280**	0.616	5.6	ug/L	48	Standard
>	**Ga**	71	135.0	44.9				mg/L	90	Standard
	Rb	85	21460.2	1.4				ug/L	23	Standard
	Y	89	449502.6	1.2				ug/L	448510	Standard
>	**Rh**	103	240.0	18.2				ug/L	15	Standard
	Mo	98	71.6	8.8	**0.0080**	0.002	28.5	ug/L	11	Standard
	Ag	107	77.3	9.2	**-0.0044**	0.001	24.3	ug/L	134	Standard
	Cd	111	102.2	7.3	**0.0460**	0.004	8.2	mg/L	7	Standard
	Cd	114	276.1	15.2	**0.0494**	0.008	16.6	ug/L	16	Standard
>	**In**	115	597587.3	0.9				ug/L	611117	Standard
	Sn	118	641.7	5.5	**0.0090**	0.005	57.4	ug/L	518	Standard
	Sb	123	117.8	11.7	**0.0127**	0.003	24.6	ug/L	127	Standard
	Ba	135	1219381.1	1.8	**603.9591**	5.851	1.0	ug/L	45	Standard
	Ce	140	360.0	14.7				ug/L	298	Standard
>	**Tb**	159	948906.4	1.3				ug/L	977942	Standard
	Ho	165	16.7	17.3				ug/L	12	Standard
	Tl	203	82.3	3.1	**0.0057**	0.000	3.2	ug/L	31	Standard
	Tl	205	61.7	24.8	**0.0067**	0.002	33.3	ug/L	20	Standard
	Pb	206	418.0	3.1	**0.0190**	0.003	17.6	ug/L	286	Standard
	Pb	207	378.0	5.3	**0.0247**	0.003	13.2	ug/L	257	Standard
	Pb	208	1532.7	1.1	**0.0263**	0.002	6.1	ug/L	956	Standard
	U	238	34.3	24.8	**0.0022**	0.001	61.6	ug/L	17	Standard
>	**Bi**	209	521633.3	1.5				ug/L	557080	Standard

Sample ID: L1606082304
Report Date/Time: Tuesday, June 28, 2016 20:57:38
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**19.9869**	34.610	173.2	mg/L	0	Standard
Mg	24	52154.6	2.1	**63.9697**	1.039	1.6	mg/L	32	Standard
K	39	548.3	16.2	**1.5625**	0.223	14.3	mg/L	30	Standard
Ca	43	246.7	5.1	**29.2175**	2.054	7.0	mg/L	48	Standard
Fe	54	45285.0	1.2	**37.6782**	1.468	3.9	mg/L	159	Standard
Fe	57	11722.9	3.7	**35.8594**	0.146	0.4	mg/L	195	Standard
> **Sc-1**	45	42894.0	3.7				mg/L	39543	Standard
Cl	35	35015.1	2.5				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	123689.6	1.5				ug/L	1647	Standard
P	31	28713.9	1.8				ug/L	15231	Standard
S	34	2633.6	10.4				ug/L	1845	Standard
Sr	88	703.3	3.4				ug/L	70	Standard
C	12	210.0	12.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	32.1	16.4				mg/L	12	Standard
Ho-1	165	16.7	17.3				mg/L	12	Standard
Er	166	26.7	57.3				mg/L	20	Standard
I	127	194264.8	2.9				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		91.091	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		99.094	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606082304
Report Date/Time: Tuesday, June 28, 2016 20:57:38
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	97.786	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	93.637	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606082304
Report Date/Time: Tuesday, June 28, 2016 20:57:38
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606084302
Sample Date/Time: Tuesday, June 28, 2016 20:58:33
Number of Replicates: 3
Autosampler Position: 320
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77271.6	2.1				ug/L	81760	Standard
	Be	9	30.0	44.1	**0.0221**	0.014	61.6	ug/L	17	Standard
	Al	27	13023291.9	1.0	**89.3957**	1.109	1.2	ug/L	1410	Standard
	Sc	45	41354.7	3.2				ug/L	39543	Standard
	Ti	47	172.3	20.3	**0.4975**	0.149	29.9	ug/L	39	Standard
	V	51	1869.8	10.2	**0.0496**	0.023	45.4	ug/L	1415	Standard
	Cr	52	9029.4	5.3	**0.1789**	0.049	27.5	ug/L	6946	Standard
	Cr	53	1933.5	16.3	**1.7631**	0.385	21.8	ug/L	328	Standard
	Mn	55	2010517.0	2.2	**295.6150**	12.044	4.1	ug/L	823	Standard
	Co	59	28381.9	0.9	**3.8204**	0.101	2.6	ug/L	252	Standard
	Ni	60	51973.3	1.5	**26.2049**	0.504	1.9	ug/L	192	Standard
	Cu	65	2375.5	2.3	**1.1012**	0.038	3.5	ug/L	141	Standard
	Zn	66	19572.0	0.3	**16.1981**	0.371	2.3	ug/L	491	Standard
>	**Ge**	72	540385.6	2.0				ug/L	515087	Standard
	As	75	349.5	8.0	**0.3364**	0.018	5.2	ug/L	20	Standard
	Se	82	126.5	4.8	**0.9278**	0.047	5.1	ug/L	15	Standard
	Se-1	77	166.7	13.3	**1.3933**	0.317	22.7	ug/L	48	Standard
>	**Ga**	71	106.7	22.2				mg/L	90	Standard
	Rb	85	40220.0	0.6				ug/L	23	Standard
	Y	89	449597.2	1.4				ug/L	448510	Standard
>	**Rh**	103	146.7	12.0				ug/L	15	Standard
	Mo	98	594.8	3.0	**0.2021**	0.008	3.8	ug/L	11	Standard
	Ag	107	92.7	31.3	**-0.0024**	0.004	183.9	ug/L	134	Standard
	Cd	111	161.3	7.4	**0.0729**	0.005	7.5	mg/L	7	Standard
	Cd	114	441.9	5.3	**0.0806**	0.004	5.2	ug/L	16	Standard
>	**In**	115	608035.7	0.5				ug/L	611117	Standard
	Sn	118	676.7	9.6	**0.0131**	0.012	88.5	ug/L	518	Standard
	Sb	123	188.3	22.2	**0.0273**	0.009	32.4	ug/L	127	Standard
	Ba	135	24154.9	0.6	**11.7345**	0.088	0.8	ug/L	45	Standard
	Ce	140	2120.1	7.8				ug/L	298	Standard
>	**Tb**	159	946861.5	2.0				ug/L	977942	Standard
	Ho	165	63.3	29.9				ug/L	12	Standard
	Tl	203	1826.8	0.7	**0.2203**	0.002	1.1	ug/L	31	Standard
	Tl	205	1685.1	3.3	**0.2367**	0.005	2.2	ug/L	20	Standard
	Pb	206	439.0	7.7	**0.0207**	0.006	26.8	ug/L	286	Standard
	Pb	207	381.0	5.7	**0.0231**	0.004	17.6	ug/L	257	Standard
	Pb	208	1509.0	1.3	**0.0226**	0.000	0.8	ug/L	956	Standard
	U	238	1189.4	4.4	**0.1693**	0.007	4.4	ug/L	17	Standard
>	**Bi**	209	536700.0	1.2				ug/L	557080	Standard

Sample ID: L1606084302
Report Date/Time: Tuesday, June 28, 2016 21:00:50
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**30.9676**	31.632	102.1	mg/L	0	Standard
Mg	24	21311.6	1.7	**27.0900**	1.076	4.0	mg/L	32	Standard
K	39	978.4	4.6	**3.0186**	0.057	1.9	mg/L	30	Standard
Ca	43	386.7	7.5	**50.8608**	4.765	9.4	mg/L	48	Standard
Fe	54	204.3	23.4	**0.0400**	0.042	106.1	mg/L	159	Standard
Fe	57	446.7	11.3	**0.8216**	0.208	25.3	mg/L	195	Standard
> **Sc-1**	45	41354.7	3.2				mg/L	39543	Standard
Cl	35	48993.8	3.0				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	7075.0	5.6				ug/L	1647	Standard
P	31	18566.4	1.9				ug/L	15231	Standard
S	34	2263.5	0.1				ug/L	1845	Standard
Sr	88	423.3	3.6				ug/L	70	Standard
C	12	123.3	40.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	104.0	29.5				mg/L	12	Standard
Ho-1	165	63.3	29.9				mg/L	12	Standard
Er	166	56.7	27.0				mg/L	20	Standard
I	127	48509.5	7.5				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.510	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		104.912	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606084302
Report Date/Time: Tuesday, June 28, 2016 21:00:50
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.496
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.342
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606084302
Report Date/Time: Tuesday, June 28, 2016 21:00:50
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 21:01:46
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77410.6	1.1				ug/L	81760	Standard
	Be	9	48694.9	4.0	**51.7876**	2.330	4.5	ug/L	17	Standard
	Al	27	7498260.2	2.7	**51.3685**	1.536	3.0	ug/L	1410	Standard
	Sc	45	39994.4	1.6				ug/L	39543	Standard
	Ti	47	24919.8	1.5	**100.1235**	0.955	1.0	ug/L	39	Standard
	V	51	352037.7	1.5	**50.4818**	0.410	0.8	ug/L	1415	Standard
	Cr	52	359056.3	3.2	**50.5608**	1.258	2.5	ug/L	6946	Standard
	Cr	53	45264.2	3.2	**51.2590**	1.232	2.4	ug/L	328	Standard
	Mn	55	341978.2	1.6	**52.6439**	0.559	1.1	ug/L	823	Standard
	Co	59	358967.4	0.9	**51.1022**	0.460	0.9	ug/L	252	Standard
	Ni	60	95904.1	3.3	**50.8527**	1.284	2.5	ug/L	192	Standard
	Cu	65	96846.7	3.1	**50.9133**	1.132	2.2	ug/L	141	Standard
	Zn	66	56366.4	2.4	**50.6359**	0.804	1.6	ug/L	491	Standard
>	**Ge**	72	514670.0	0.9				ug/L	515087	Standard
	As	75	56653.0	2.7	**50.4646**	0.946	1.9	ug/L	20	Standard
	Se	82	5738.6	2.7	**49.7611**	0.957	1.9	ug/L	15	Standard
	Se-1	77	3922.2	4.5	**51.3791**	2.097	4.1	ug/L	48	Standard
>	**Ga**	71	78.3	3.7				mg/L	90	Standard
	Rb	85	913.4	7.9				ug/L	23	Standard
	Y	89	449629.1	4.0				ug/L	448510	Standard
>	**Rh**	103	26.7	60.3				ug/L	15	Standard
	Mo	98	269192.0	0.5	**99.9193**	1.745	1.7	ug/L	11	Standard
	Ag	107	324356.7	1.6	**48.4664**	0.560	1.2	ug/L	134	Standard
	Cd	111	104800.0	2.8	**49.1868**	1.080	2.2	mg/L	7	Standard
	Cd	114	253765.7	3.7	**48.9997**	1.658	3.4	ug/L	16	Standard
>	**In**	115	609141.6	2.1				ug/L	611117	Standard
	Sn	118	292076.9	3.3	**49.4485**	0.824	1.7	ug/L	518	Standard
	Sb	123	232366.9	1.6	**49.3998**	0.394	0.8	ug/L	127	Standard
	Ba	135	100964.1	2.8	**49.0412**	1.097	2.2	ug/L	45	Standard
	Ce	140	81.7	25.5				ug/L	298	Standard
>	**Tb**	159	939122.1	2.3				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	411985.1	1.5	**50.6376**	0.301	0.6	ug/L	31	Standard
	Tl	205	371737.4	2.9	**52.6085**	0.609	1.2	ug/L	20	Standard
	Pb	206	267607.7	0.9	**51.2233**	0.555	1.1	ug/L	286	Standard
	Pb	207	243032.3	0.6	**51.5800**	0.780	1.5	ug/L	257	Standard
	Pb	208	974203.2	1.1	**53.4395**	0.502	0.9	ug/L	956	Standard
	U	238	359093.2	2.3	**51.8989**	0.179	0.3	ug/L	17	Standard
>	**Bi**	209	537912.3	2.0				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:04:03
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3960.5	5.8	**5.1535**	0.239	4.6	mg/L	32	Standard
K	39	1450.1	5.1	**4.7011**	0.180	3.8	mg/L	30	Standard
Ca	43	56.7	22.2	**3.1676**	1.896	59.9	mg/L	48	Standard
Fe	54	5602.2	1.3	**4.8773**	0.145	3.0	mg/L	159	Standard
Fe	57	1671.8	3.1	**4.9554**	0.149	3.0	mg/L	195	Standard
Sc-1	45	39994.4	1.6				mg/L	39543	Standard
Cl	35	34924.3	5.3				ug/L	27642	Standard
Kr	83	2.3	49.5				ug/L	2	Standard
Br	81	3010.3	3.5				ug/L	1647	Standard
P	31	18342.8	3.8				ug/L	15231	Standard
S	34	2438.5	4.7				ug/L	1845	Standard
Sr	88	55.0	47.2				ug/L	70	Standard
C	12	96.7	23.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	23.2	100.3				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	3.3	173.2				mg/L	20	Standard
I	127	10668.8	10.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	103.575		
Al	27	102.737		
Sc	45			
Ti	47	100.123		
V	51	100.964		
Cr	52	101.122		
Cr	53			
Mn	55	105.288		
Co	59	102.204		
Ni	60	101.705		
Cu	65	101.827		
Zn	66	101.272		
Ge	72		99.919	
As	75	100.929		
Se	82	99.522		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:04:03
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	99.919		
Ag	107	96.933		
Cd	111	98.374		
Cd	114			
> In	115			99.677
Sn	118	98.897		
Sb	123	98.800		
Ba	135	98.082		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.275		
Tl	205			
Pb	206			
Pb	207			
Pb	208	106.879		
U	238	103.798		
> Bi	209			96.559
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:04:03
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 21:04:57
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76325.1	1.4				ug/L	81760	Standard
	Be	9	36.7	47.9	**0.0297**	0.018	61.9	ug/L	17	Standard
	Al	27	7433.8	52.1	**0.0370**	0.026	70.5	ug/L	1410	Standard
	Sc	45	39083.7	0.3				ug/L	39543	Standard
	Ti	47	29.0	21.5	**-0.0509**	0.024	47.8	ug/L	39	Standard
	V	51	1130.0	4.7	**-0.0473**	0.008	16.3	ug/L	1415	Standard
	Cr	52	6124.9	1.6	**-0.1950**	0.012	6.3	ug/L	6946	Standard
	Cr	53	786.7	2.0	**0.5367**	0.019	3.5	ug/L	328	Standard
	Mn	55	1151.0	22.6	**0.0255**	0.040	155.2	ug/L	823	Standard
	Co	59	234.7	8.8	**0.0014**	0.003	211.9	ug/L	252	Standard
	Ni	60	158.3	5.7	**-0.0226**	0.005	21.5	ug/L	192	Standard
	Cu	65	164.3	12.7	**-0.0061**	0.011	179.5	ug/L	141	Standard
	Zn	66	470.3	2.4	**-0.4072**	0.010	2.5	ug/L	491	Standard
>	**Ge**	72	525676.3	0.2				ug/L	515087	Standard
	As	75	-22.3	21.6	**0.0209**	0.004	19.9	ug/L	20	Standard
	Se	82	22.8	31.8	**0.0747**	0.062	82.4	ug/L	15	Standard
	Se-1	77	66.0	18.9	**0.1398**	0.164	117.4	ug/L	48	Standard
>	**Ga**	71	56.7	10.2				mg/L	90	Standard
	Rb	85	51.7	11.2				ug/L	23	Standard
	Y	89	445574.4	3.2				ug/L	448510	Standard
>	**Rh**	103	28.3	36.7				ug/L	15	Standard
	Mo	98	286.8	35.9	**0.0884**	0.038	42.5	ug/L	11	Standard
	Ag	107	119.3	19.2	**0.0018**	0.003	188.2	ug/L	134	Standard
	Cd	111	12.1	17.7	**0.0028**	0.001	35.0	mg/L	7	Standard
	Cd	114	42.6	30.2	**0.0034**	0.003	75.4	ug/L	16	Standard
>	**In**	115	602370.1	0.9				ug/L	611117	Standard
	Sn	118	773.4	17.1	**0.0309**	0.024	76.6	ug/L	518	Standard
	Sb	123	433.3	25.4	**0.0805**	0.024	30.2	ug/L	127	Standard
	Ba	135	290.7	43.2	**0.1174**	0.060	51.4	ug/L	45	Standard
	Ce	140	18.3	56.8				ug/L	298	Standard
>	**Tb**	159	950012.1	0.7				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	71.7	40.5	**0.0037**	0.004	95.1	ug/L	31	Standard
	Tl	205	85.0	56.1	**0.0094**	0.007	71.5	ug/L	20	Standard
	Pb	206	271.0	4.2	**-0.0136**	0.003	20.1	ug/L	286	Standard
	Pb	207	260.3	8.3	**-0.0047**	0.005	107.3	ug/L	257	Standard
	Pb	208	1029.7	7.7	**-0.0060**	0.005	82.3	ug/L	956	Standard
	U	238	47.7	42.8	**0.0037**	0.003	79.6	ug/L	17	Standard
>	**Bi**	209	558519.2	1.3				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:07:14
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	48.3	11.9	**0.0079**	0.008	95.2	mg/L	32	Standard
K	39	18.3	56.8	**-0.0788**	0.036	45.2	mg/L	30	Standard
Ca	43	31.7	24.1	**-0.4803**	1.172	244.0	mg/L	48	Standard
Fe	54	125.9	16.2	**-0.0218**	0.019	87.4	mg/L	159	Standard
Fe	57	238.3	12.6	**0.1894**	0.102	54.0	mg/L	195	Standard
> **Sc-1**	45	39083.7	0.3				mg/L	39543	Standard
Cl	35	33723.5	2.9				ug/L	27642	Standard
Kr	83	2.0	100.0				ug/L	2	Standard
Br	81	2386.9	5.5				ug/L	1647	Standard
P	31	18112.5	4.7				ug/L	15231	Standard
S	34	2391.9	4.0				ug/L	1845	Standard
Sr	88	68.3	29.6				ug/L	70	Standard
C	12	100.0	36.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	18.9	54.4				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	23.3	65.5				mg/L	20	Standard
I	127	7351.8	8.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.056	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:07:14
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	98.569	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	100.258	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:07:14
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141001
Sample Date/Time: Tuesday, June 28, 2016 21:08:10
Number of Replicates: 3
Autosampler Position: 321
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79030.6	2.3				ug/L	81760	Standard
	Be	9	11.7	89.2	**0.0026**	0.011	425.8	ug/L	17	Standard
	Al	27	44776805.2	0.7	**300.5948**	6.187	2.1	ug/L	1410	Standard
	Sc	45	43343.6	2.1				ug/L	39543	Standard
	Ti	47	192.3	8.0	**0.6107**	0.056	9.2	ug/L	39	Standard
	V	51	4772.2	4.6	**0.4823**	0.029	6.1	ug/L	1415	Standard
	Cr	52	9840.2	0.5	**0.3614**	0.015	4.2	ug/L	6946	Standard
	Cr	53	1230.1	5.5	**1.0645**	0.072	6.7	ug/L	328	Standard
	Mn	55	4769620.0	0.5	**738.1503**	9.110	1.2	ug/L	823	Standard
	Co	59	9207.1	1.6	**1.2837**	0.028	2.2	ug/L	252	Standard
	Ni	60	5238.9	2.1	**2.6865**	0.062	2.3	ug/L	192	Standard
	Cu	65	718.7	4.1	**0.2886**	0.013	4.4	ug/L	141	Standard
	Zn	66	2924.3	0.9	**1.8496**	0.029	1.6	ug/L	491	Standard
>	**Ge**	72	513321.3	0.8				ug/L	515087	Standard
	As	75	4203.4	1.1	**3.7919**	0.050	1.3	ug/L	20	Standard
	Se	82	127.5	8.3	**0.9927**	0.098	9.9	ug/L	15	Standard
	Se-1	77	88.3	31.6	**0.4581**	0.375	81.9	ug/L	48	Standard
>	**Ga**	71	145.0	24.1				mg/L	90	Standard
	Rb	85	2930.3	3.3				ug/L	23	Standard
	Y	89	443412.2	1.4				ug/L	448510	Standard
>	**Rh**	103	140.0	15.6				ug/L	15	Standard
	Mo	98	1185.9	4.8	**0.4170**	0.022	5.4	ug/L	11	Standard
	Ag	107	98.0	15.7	**-0.0017**	0.002	128.2	ug/L	134	Standard
	Cd	111	19.9	13.4	**0.0063**	0.001	18.0	mg/L	7	Standard
	Cd	114	87.1	38.1	**0.0117**	0.006	53.6	ug/L	16	Standard
>	**In**	115	614969.4	1.2				ug/L	611117	Standard
	Sn	118	790.0	17.2	**0.0309**	0.023	75.8	ug/L	518	Standard
	Sb	123	280.0	20.4	**0.0462**	0.012	26.5	ug/L	127	Standard
	Ba	135	192282.6	1.4	**92.5375**	1.689	1.8	ug/L	45	Standard
	Ce	140	2006.8	4.1				ug/L	298	Standard
>	**Tb**	159	961897.3	2.5				ug/L	977942	Standard
	Ho	165	81.7	52.8				ug/L	12	Standard
	Tl	203	119.7	8.5	**0.0099**	0.002	15.4	ug/L	31	Standard
	Tl	205	98.3	15.5	**0.0116**	0.002	19.8	ug/L	20	Standard
	Pb	206	482.7	4.9	**0.0287**	0.002	8.4	ug/L	286	Standard
	Pb	207	424.0	4.8	**0.0319**	0.005	14.5	ug/L	257	Standard
	Pb	208	1694.0	1.8	**0.0325**	0.004	11.5	ug/L	956	Standard
	U	238	15563.7	1.1	**2.2446**	0.071	3.2	ug/L	17	Standard
>	**Bi**	209	538649.3	2.3				ug/L	557080	Standard

Sample ID: L1606141001
Report Date/Time: Tuesday, June 28, 2016 21:10:27
Page 1

Approved: June 29, 2016

Na	23	16.7	75.5	**98.5166**	75.230	76.4	mg/L	0	Standard
Mg	24	9014.3	3.2	**10.8981**	0.573	5.3	mg/L	32	Standard
K	39	116.7	19.3	**0.2184**	0.072	32.8	mg/L	30	Standard
Ca	43	525.0	6.7	**67.4527**	5.707	8.5	mg/L	48	Standard
Fe	54	1820.8	2.9	**1.3668**	0.058	4.2	mg/L	159	Standard
Fe	57	973.4	7.5	**2.3749**	0.248	10.4	mg/L	195	Standard
> **Sc-1**	45	43343.6	2.1				mg/L	39543	Standard
Cl	35	38331.8	2.6				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	27825.5	3.4				ug/L	1647	Standard
P	31	23363.0	0.7				ug/L	15231	Standard
S	34	2495.2	2.5				ug/L	1845	Standard
Sr	88	533.3	5.7				ug/L	70	Standard
C	12	353.3	5.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	83.6	23.1				mg/L	12	Standard
Ho-1	165	81.7	52.8				mg/L	12	Standard
Er	166	63.3	50.8				mg/L	20	Standard
I	127	532195.8	11.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		96.662	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.657	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606141001
Report Date/Time: Tuesday, June 28, 2016 21:10:27
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	100.630
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	96.691
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141201
Sample Date/Time: Tuesday, June 28, 2016 21:11:22
Number of Replicates: 3
Autosampler Position: 322
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79318.7	0.7				ug/L	81760	Standard
	Be	9	16.7	45.8	**0.0076**	0.008	106.2	ug/L	17	Standard
	Al	27	14913651.7	1.4	**99.7256**	2.143	2.1	ug/L	1410	Standard
	Sc	45	40395.5	2.6				ug/L	39543	Standard
	Ti	47	500.3	4.7	**1.8979**	0.055	2.9	ug/L	39	Standard
	V	51	36055.9	2.6	**5.1139**	0.040	0.8	ug/L	1415	Standard
	Cr	52	13456.3	2.4	**0.9269**	0.086	9.3	ug/L	6946	Standard
	Cr	53	9844.9	11.5	**11.1480**	1.101	9.9	ug/L	328	Standard
	Mn	55	27340.3	1.0	**4.1779**	0.048	1.1	ug/L	823	Standard
	Co	59	924.7	1.8	**0.1037**	0.005	4.9	ug/L	252	Standard
	Ni	60	2165.5	2.9	**1.0744**	0.025	2.4	ug/L	192	Standard
	Cu	65	1308.7	4.1	**0.6156**	0.023	3.7	ug/L	141	Standard
	Zn	66	7243.7	2.3	**5.9517**	0.120	2.0	ug/L	491	Standard
>	**Ge**	72	502171.0	2.0				ug/L	515087	Standard
	As	75	1256.7	3.1	**1.1867**	0.017	1.5	ug/L	20	Standard
	Se	82	736.7	2.2	**6.4443**	0.037	0.6	ug/L	15	Standard
	Se-1	77	926.0	3.3	**11.8887**	0.285	2.4	ug/L	48	Standard
>	**Ga**	71	176.7	6.5				mg/L	90	Standard
	Rb	85	2561.9	8.5				ug/L	23	Standard
	Y	89	439495.0	0.9				ug/L	448510	Standard
>	**Rh**	103	90.0	38.9				ug/L	15	Standard
	Mo	98	758.7	2.1	**0.2744**	0.005	1.7	ug/L	11	Standard
	Ag	107	112.7	8.8	**0.0013**	0.001	110.1	ug/L	134	Standard
	Cd	111	18.4	13.8	**0.0061**	0.001	20.0	mg/L	7	Standard
	Cd	114	55.1	29.0	**0.0062**	0.003	52.0	ug/L	16	Standard
>	**In**	115	584464.7	0.6				ug/L	611117	Standard
	Sn	118	741.7	7.5	**0.0292**	0.010	34.2	ug/L	518	Standard
	Sb	123	322.5	19.3	**0.0586**	0.014	23.6	ug/L	127	Standard
	Ba	135	23648.1	1.5	**11.9524**	0.216	1.8	ug/L	45	Standard
	Ce	140	4514.0	4.6				ug/L	298	Standard
>	**Tb**	159	922323.8	0.6				ug/L	977942	Standard
	Ho	165	91.7	12.6				ug/L	12	Standard
	Tl	203	130.7	10.2	**0.0124**	0.002	15.2	ug/L	31	Standard
	Tl	205	113.3	46.8	**0.0149**	0.008	53.9	ug/L	20	Standard
	Pb	206	585.3	3.6	**0.0562**	0.004	7.1	ug/L	286	Standard
	Pb	207	533.0	4.2	**0.0629**	0.005	7.8	ug/L	257	Standard
	Pb	208	2107.4	5.4	**0.0632**	0.008	12.1	ug/L	956	Standard
	U	238	32845.9	1.3	**5.0682**	0.050	1.0	ug/L	17	Standard
>	**Bi**	209	503585.9	0.8				ug/L	557080	Standard

Sample ID: L1606141201
Report Date/Time: Tuesday, June 28, 2016 21:13:39
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	118021.2	0.8	**153.7808**	2.811	1.8	mg/L	32	Standard
K	39	101.7	17.3	**0.1948**	0.058	29.7	mg/L	30	Standard
Ca	43	335.0	15.7	**44.3505**	6.519	14.7	mg/L	48	Standard
Fe	54	237.2	9.3	**0.0729**	0.016	22.5	mg/L	159	Standard
Fe	57	498.3	12.0	**1.0226**	0.198	19.4	mg/L	195	Standard
> **Sc-1**	45	40395.5	2.6				mg/L	39543	Standard
Cl	35	43781.2	1.0				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	58978.2	6.0				ug/L	1647	Standard
P	31	22017.7	1.7				ug/L	15231	Standard
S	34	2351.8	1.8				ug/L	1845	Standard
Sr	88	240.0	11.0				ug/L	70	Standard
C	12	356.7	26.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	169.3	7.0				mg/L	12	Standard
Ho-1	165	91.7	12.6				mg/L	12	Standard
Er	166	83.3	13.9				mg/L	20	Standard
I	127	31387.7	30.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		97.014	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.492	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606141201
Report Date/Time: Tuesday, June 28, 2016 21:13:39
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.639
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		90.397
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141202
Sample Date/Time: Tuesday, June 28, 2016 21:14:33
Number of Replicates: 3
Autosampler Position: 323
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83650.3	1.0				ug/L	81760	Standard
	Be	9	21.7	53.3	**0.0115**	0.011	97.0	ug/L	17	Standard
	Al	27	38632215.5	2.3	**244.9578**	5.868	2.4	ug/L	1410	Standard
	Sc	45	43420.5	1.6				ug/L	39543	Standard
	Ti	47	186.0	10.8	**0.5354**	0.077	14.4	ug/L	39	Standard
	V	51	26880.1	4.2	**3.4128**	0.061	1.8	ug/L	1415	Standard
	Cr	52	10918.6	2.1	**0.4115**	0.017	4.1	ug/L	6946	Standard
	Cr	53	7481.8	4.1	**7.6462**	0.526	6.9	ug/L	328	Standard
	Mn	55	33664.4	1.6	**4.7149**	0.144	3.1	ug/L	823	Standard
	Co	59	762.7	5.1	**0.0704**	0.007	10.1	ug/L	252	Standard
	Ni	60	2375.2	1.6	**1.0759**	0.011	1.0	ug/L	192	Standard
	Cu	65	711.7	7.3	**0.2600**	0.027	10.4	ug/L	141	Standard
	Zn	66	3134.0	1.7	**1.8510**	0.112	6.1	ug/L	491	Standard
>	**Ge**	72	550214.2	2.6				ug/L	515087	Standard
	As	75	794.1	8.2	**0.7027**	0.072	10.2	ug/L	20	Standard
	Se	82	234.2	3.9	**1.7863**	0.091	5.1	ug/L	15	Standard
	Se-1	77	482.0	4.4	**5.2700**	0.176	3.3	ug/L	48	Standard
>	**Ga**	71	80.0	16.5				mg/L	90	Standard
	Rb	85	1268.4	2.0				ug/L	23	Standard
	Y	89	467702.3	2.0				ug/L	448510	Standard
>	**Rh**	103	125.0	4.0				ug/L	15	Standard
	Mo	98	395.0	8.4	**0.1254**	0.010	8.1	ug/L	11	Standard
	Ag	107	91.7	6.0	**-0.0027**	0.001	31.4	ug/L	134	Standard
	Cd	111	17.7	18.1	**0.0053**	0.002	30.5	mg/L	7	Standard
	Cd	114	70.4	32.3	**0.0084**	0.004	47.5	ug/L	16	Standard
>	**In**	115	618005.7	2.3				ug/L	611117	Standard
	Sn	118	668.3	14.3	**0.0101**	0.018	180.4	ug/L	518	Standard
	Sb	123	172.5	13.4	**0.0233**	0.005	22.3	ug/L	127	Standard
	Ba	135	80657.5	2.0	**38.6136**	0.662	1.7	ug/L	45	Standard
	Ce	140	231.7	23.7				ug/L	298	Standard
>	**Tb**	159	976289.5	2.7				ug/L	977942	Standard
	Ho	165	20.0	50.0				ug/L	12	Standard
	Tl	203	130.7	0.4	**0.0112**	0.000	4.3	ug/L	31	Standard
	Tl	205	103.3	15.6	**0.0123**	0.003	20.6	ug/L	20	Standard
	Pb	206	343.3	3.7	**0.0019**	0.004	206.6	ug/L	286	Standard
	Pb	207	309.0	3.8	**0.0073**	0.004	55.0	ug/L	257	Standard
	Pb	208	1267.4	6.3	**0.0088**	0.006	66.3	ug/L	956	Standard
	U	238	23437.4	1.1	**3.3661**	0.068	2.0	ug/L	17	Standard
>	**Bi**	209	541066.6	2.7				ug/L	557080	Standard

Sample ID: L1606141202
Report Date/Time: Tuesday, June 28, 2016 21:16:50
Page 1

Approved: June 29, 2016

Na	23	10.0	100.0	**57.8886**	57.492	99.3	mg/L	0	Standard
Mg	24	42571.4	0.8	**51.5637**	1.207	2.3	mg/L	32	Standard
K	39	90.0	30.9	**0.1357**	0.087	64.2	mg/L	30	Standard
Ca	43	438.3	26.2	**55.1842**	15.190	27.5	mg/L	48	Standard
Fe	54	177.1	8.1	**0.0090**	0.012	139.0	mg/L	159	Standard
Fe	57	461.7	5.0	**0.7949**	0.071	8.9	mg/L	195	Standard
Sc-1	45	43420.5	1.6				mg/L	39543	Standard
Cl	35	42170.0	2.1				ug/L	27642	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	34382.4	5.6				ug/L	1647	Standard
P	31	23767.0	4.2				ug/L	15231	Standard
S	34	2608.6	3.1				ug/L	1845	Standard
Sr	88	431.7	5.4				ug/L	70	Standard
C	12	273.3	10.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	32.2	46.2				mg/L	12	Standard
Ho-1	165	20.0	50.0				mg/L	12	Standard
Er	166	23.3	89.2				mg/L	20	Standard
I	127	35633.6	4.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		102.312	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		106.820	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606141202
Report Date/Time: Tuesday, June 28, 2016 21:16:50
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	101.127	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	97.125	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606141202
Report Date/Time: Tuesday, June 28, 2016 21:16:50
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141301
Sample Date/Time: Tuesday, June 28, 2016 21:17:45
Number of Replicates: 3
Autosampler Position: 324
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79916.9	1.5				ug/L	81760	Standard
	Be	9	28.3	27.0	**0.0194**	0.007	38.6	ug/L	17	Standard
	Al	27	38023288.5	2.1	**252.3385**	1.857	0.7	ug/L	1410	Standard
	Sc	45	42571.5	4.4				ug/L	39543	Standard
	Ti	47	396.7	7.8	**1.4239**	0.097	6.8	ug/L	39	Standard
	V	51	6119.0	6.6	**0.6719**	0.073	10.9	ug/L	1415	Standard
	Cr	52	11441.3	0.3	**0.5818**	0.031	5.4	ug/L	6946	Standard
	Cr	53	9654.8	11.4	**10.6403**	1.449	13.6	ug/L	328	Standard
	Mn	55	1634887.0	2.2	**251.2478**	2.194	0.9	ug/L	823	Standard
	Co	59	3215.3	1.8	**0.4249**	0.007	1.5	ug/L	252	Standard
	Ni	60	4170.9	0.8	**2.1033**	0.039	1.9	ug/L	192	Standard
	Cu	65	1200.4	3.8	**0.5391**	0.026	4.9	ug/L	141	Standard
	Zn	66	3234.7	1.4	**2.1147**	0.035	1.7	ug/L	491	Standard
>	**Ge**	72	516685.2	1.8				ug/L	515087	Standard
	As	75	16545.6	0.8	**14.7132**	0.205	1.4	ug/L	20	Standard
	Se	82	192.8	5.0	**1.5497**	0.062	4.0	ug/L	15	Standard
	Se-1	77	648.0	6.7	**7.8638**	0.716	9.1	ug/L	48	Standard
>	**Ga**	71	125.0	18.3				mg/L	90	Standard
	Rb	85	1723.4	7.4				ug/L	23	Standard
	Y	89	425819.3	3.1				ug/L	448510	Standard
>	**Rh**	103	118.3	17.1				ug/L	15	Standard
	Mo	98	4050.2	1.8	**1.5330**	0.030	2.0	ug/L	11	Standard
	Ag	107	94.7	12.2	**-0.0016**	0.002	127.1	ug/L	134	Standard
	Cd	111	117.7	9.5	**0.0542**	0.007	12.3	mg/L	7	Standard
	Cd	114	297.9	4.3	**0.0546**	0.005	8.5	ug/L	16	Standard
>	**In**	115	590223.7	3.4				ug/L	611117	Standard
	Sn	118	1010.0	10.9	**0.0746**	0.013	17.7	ug/L	518	Standard
	Sb	123	1963.6	1.5	**0.4183**	0.013	3.0	ug/L	127	Standard
	Ba	135	194231.1	1.1	**97.4394**	2.308	2.4	ug/L	45	Standard
	Ce	140	3857.2	6.5				ug/L	298	Standard
>	**Tb**	159	963155.2	4.7				ug/L	977942	Standard
	Ho	165	81.7	25.5				ug/L	12	Standard
	Tl	203	50.3	12.9	**0.0015**	0.001	66.6	ug/L	31	Standard
	Tl	205	46.7	32.7	**0.0044**	0.002	52.1	ug/L	20	Standard
	Pb	206	2455.5	1.0	**0.4110**	0.018	4.4	ug/L	286	Standard
	Pb	207	2061.1	1.1	**0.3836**	0.016	4.1	ug/L	257	Standard
	Pb	208	8494.6	1.5	**0.4098**	0.011	2.7	ug/L	956	Standard
	U	238	9671.1	1.3	**1.4065**	0.038	2.7	ug/L	17	Standard
>	**Bi**	209	533828.4	3.8				ug/L	557080	Standard

Sample ID: L1606141301
Report Date/Time: Tuesday, June 28, 2016 21:20:02
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	104970.9	3.0	**129.8006**	2.883	2.2	mg/L	32	Standard
K	39	241.7	3.2	**0.6187**	0.058	9.4	mg/L	30	Standard
Ca	43	386.7	15.6	**49.0389**	6.282	12.8	mg/L	48	Standard
Fe	54	7303.8	3.7	**6.0094**	0.321	5.3	mg/L	159	Standard
Fe	57	2221.8	6.0	**6.3400**	0.172	2.7	mg/L	195	Standard
> **Sc-1**	45	42571.5	4.4				mg/L	39543	Standard
Cl	35	37833.2	3.5				ug/L	27642	Standard
Kr	83	4.7	44.6				ug/L	2	Standard
Br	81	42533.0	4.1				ug/L	1647	Standard
P	31	23852.1	1.0				ug/L	15231	Standard
S	34	2610.2	4.2				ug/L	1845	Standard
Sr	88	403.3	5.0				ug/L	70	Standard
C	12	1170.0	10.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	86.6				mg/L	0	Standard
Dy	164	115.7	9.7				mg/L	12	Standard
Ho-1	165	81.7	25.5				mg/L	12	Standard
Er	166	90.0	29.4				mg/L	20	Standard
I	127	137002.4	10.0				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		97.746	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		100.310	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606141301
Report Date/Time: Tuesday, June 28, 2016 21:20:02
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.581
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.826
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606141301
Report Date/Time: Tuesday, June 28, 2016 21:20:02
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141601
Sample Date/Time: Tuesday, June 28, 2016 21:20:56
Number of Replicates: 3
Autosampler Position: 325
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92372.4	1.4				ug/L	81760	Standard
	Be	9	16.7	45.8	**0.0051**	0.007	132.4	ug/L	17	Standard
	Al	27	74980741.0	2.1	**430.5985**	11.194	2.6	ug/L	1410	Standard
	Sc	45	42273.9	1.4				ug/L	39543	Standard
	Ti	47	206.0	5.5	**0.6860**	0.049	7.2	ug/L	39	Standard
	V	51	76665.0	2.1	**11.1242**	0.138	1.2	ug/L	1415	Standard
	Cr	52	20589.0	2.5	**1.9813**	0.052	2.6	ug/L	6946	Standard
	Cr	53	16427.3	7.3	**18.8858**	1.445	7.7	ug/L	328	Standard
	Mn	55	4403.6	9.9	**0.5498**	0.075	13.7	ug/L	823	Standard
	Co	59	1375.4	0.7	**0.1698**	0.002	1.0	ug/L	252	Standard
	Ni	60	4849.1	3.4	**2.5400**	0.080	3.2	ug/L	192	Standard
	Cu	65	2467.2	2.0	**1.2432**	0.032	2.6	ug/L	141	Standard
	Zn	66	3268.7	3.6	**2.2359**	0.099	4.4	ug/L	491	Standard
>	**Ge**	72	501391.0	1.2				ug/L	515087	Standard
	As	75	1723.8	1.9	**1.6153**	0.015	1.0	ug/L	20	Standard
	Se	82	1818.6	2.0	**16.1069**	0.131	0.8	ug/L	15	Standard
	Se-1	77	2238.8	2.0	**29.8099**	0.421	1.4	ug/L	48	Standard
>	**Ga**	71	70.0	35.7				mg/L	90	Standard
	Rb	85	6024.5	1.3				ug/L	23	Standard
	Y	89	451516.3	2.4				ug/L	448510	Standard
>	**Rh**	103	580.0	3.0				ug/L	15	Standard
	Mo	98	4894.1	3.5	**1.7784**	0.027	1.5	ug/L	11	Standard
	Ag	107	107.3	12.0	**-0.0004**	0.002	428.3	ug/L	134	Standard
	Cd	111	61.6	14.8	**0.0258**	0.005	19.6	mg/L	7	Standard
	Cd	114	169.0	6.0	**0.0274**	0.002	5.8	ug/L	16	Standard
>	**In**	115	615602.8	3.1				ug/L	611117	Standard
	Sn	118	783.4	1.3	**0.0296**	0.003	11.3	ug/L	518	Standard
	Sb	123	203.3	3.5	**0.0299**	0.001	4.8	ug/L	127	Standard
	Ba	135	8199.6	0.5	**3.9202**	0.126	3.2	ug/L	45	Standard
	Ce	140	360.0	9.7				ug/L	298	Standard
>	**Tb**	159	979932.9	1.3				ug/L	977942	Standard
	Ho	165	28.3	20.4				ug/L	12	Standard
	Tl	203	89.3	17.1	**0.0065**	0.002	28.6	ug/L	31	Standard
	Tl	205	80.0	25.0	**0.0093**	0.003	29.8	ug/L	20	Standard
	Pb	206	345.7	2.0	**0.0043**	0.002	47.1	ug/L	286	Standard
	Pb	207	309.7	5.2	**0.0094**	0.004	39.7	ug/L	257	Standard
	Pb	208	1290.0	0.3	**0.0121**	0.001	10.3	ug/L	956	Standard
	U	238	27903.3	2.3	**4.1288**	0.057	1.4	ug/L	17	Standard
>	**Bi**	209	525055.5	1.5				ug/L	557080	Standard

Sample ID: L1606141601
Report Date/Time: Tuesday, June 28, 2016 21:23:13
Page 1

Approved: June 29, 2016

Na	23	21.7	48.0	**130.4724**	63.085	48.4	mg/L	0	Standard
Mg	24	101248.9	2.2	**126.0185**	2.508	2.0	mg/L	32	Standard
K	39	313.3	18.1	**0.8505**	0.192	22.6	mg/L	30	Standard
Ca	43	933.4	7.3	**127.1886**	7.858	6.2	mg/L	48	Standard
Fe	54	133.4	13.8	**-0.0242**	0.015	63.9	mg/L	159	Standard
Fe	57	858.4	5.3	**2.0873**	0.173	8.3	mg/L	195	Standard
> **Sc-1**	45	42273.9	1.4				mg/L	39543	Standard
Cl	35	65275.0	2.4				ug/L	27642	Standard
Kr	83	0.7	86.6				ug/L	2	Standard
Br	81	32608.5	7.2				ug/L	1647	Standard
P	31	23807.1	4.7				ug/L	15231	Standard
S	34	2403.5	5.0				ug/L	1845	Standard
Sr	88	2410.2	4.8				ug/L	70	Standard
C	12	193.3	34.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	0	Standard
Dy	164	21.1	69.3				mg/L	12	Standard
Ho-1	165	28.3	20.4				mg/L	12	Standard
Er	166	46.7	53.9				mg/L	20	Standard
I	127	42467.9	5.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		112.980	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.341	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606141601
Report Date/Time: Tuesday, June 28, 2016 21:23:13
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.734
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.251
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606141601
Report Date/Time: Tuesday, June 28, 2016 21:23:13
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606141602
Sample Date/Time: Tuesday, June 28, 2016 21:24:08
Number of Replicates: 3
Autosampler Position: 326
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88434.1	2.2				ug/L	81760	Standard
	Be	9	13.3	57.3	**0.0028**	0.007	262.6	ug/L	17	Standard
	Al	27	52144460.9	0.4	**312.8622**	8.112	2.6	ug/L	1410	Standard
	Sc	45	37740.3	1.5				ug/L	39543	Standard
	Ti	47	227.7	10.0	**0.8076**	0.101	12.5	ug/L	39	Standard
	V	51	38489.9	1.8	**5.6761**	0.140	2.5	ug/L	1415	Standard
	Cr	52	19640.1	1.0	**1.9398**	0.010	0.5	ug/L	6946	Standard
	Cr	53	17715.4	1.9	**21.0942**	0.301	1.4	ug/L	328	Standard
	Mn	55	4373.0	6.9	**0.5679**	0.048	8.4	ug/L	823	Standard
	Co	59	870.4	1.6	**0.1002**	0.001	1.2	ug/L	252	Standard
	Ni	60	2666.6	2.1	**1.3989**	0.023	1.6	ug/L	192	Standard
	Cu	65	1074.7	4.8	**0.5100**	0.030	5.9	ug/L	141	Standard
	Zn	66	2890.6	2.0	**1.9736**	0.048	2.4	ug/L	491	Standard
>	**Ge**	72	484924.7	0.7				ug/L	515087	Standard
	As	75	2458.8	5.8	**2.3632**	0.133	5.6	ug/L	20	Standard
	Se	82	892.7	6.7	**8.1181**	0.571	7.0	ug/L	15	Standard
	Se-1	77	1757.1	0.5	**24.0557**	0.109	0.5	ug/L	48	Standard
>	**Ga**	71	76.7	13.6				mg/L	90	Standard
	Rb	85	4900.8	2.1				ug/L	23	Standard
	Y	89	417881.5	5.4				ug/L	448510	Standard
>	**Rh**	103	255.0	18.7				ug/L	15	Standard
	Mo	98	772.2	2.1	**0.2950**	0.006	2.2	ug/L	11	Standard
	Ag	107	68.7	21.6	**-0.0049**	0.003	53.5	ug/L	134	Standard
	Cd	111	15.0	33.1	**0.0048**	0.003	56.9	mg/L	7	Standard
	Cd	114	24.6	131.5	**0.0004**	0.007	1725.2	ug/L	16	Standard
>	**In**	115	555990.7	3.0				ug/L	611117	Standard
	Sn	118	696.7	10.1	**0.0278**	0.015	54.6	ug/L	518	Standard
	Sb	123	106.5	16.8	**0.0120**	0.005	38.5	ug/L	127	Standard
	Ba	135	12524.8	3.0	**6.6489**	0.344	5.2	ug/L	45	Standard
	Ce	140	808.4	5.3				ug/L	298	Standard
>	**Tb**	159	897664.5	1.7				ug/L	977942	Standard
	Ho	165	23.3	68.9				ug/L	12	Standard
	Tl	203	89.0	17.6	**0.0071**	0.002	31.3	ug/L	31	Standard
	Tl	205	90.0	0.0	**0.0115**	0.000	1.5	ug/L	20	Standard
	Pb	206	365.7	4.1	**0.0124**	0.004	32.4	ug/L	286	Standard
	Pb	207	325.3	1.4	**0.0169**	0.000	0.6	ug/L	257	Standard
	Pb	208	1289.4	1.1	**0.0163**	0.000	2.7	ug/L	956	Standard
	U	238	25222.0	1.6	**3.9519**	0.013	0.3	ug/L	17	Standard
>	**Bi**	209	495840.1	1.2				ug/L	557080	Standard

Sample ID: L1606141602
Report Date/Time: Tuesday, June 28, 2016 21:26:25
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**22.3577**	19.364	86.6	mg/L	0	Standard
Mg	24	65916.4	2.1	**91.8901**	2.205	2.4	mg/L	32	Standard
K	39	221.7	4.7	**0.6432**	0.031	4.9	mg/L	30	Standard
Ca	43	530.0	9.4	**79.0853**	9.116	11.5	mg/L	48	Standard
Fe	54	138.0	10.0	**-0.0062**	0.012	195.9	mg/L	159	Standard
Fe	57	510.0	5.1	**1.1793**	0.086	7.3	mg/L	195	Standard
> **Sc-1**	45	37740.3	1.5				mg/L	39543	Standard
Cl	35	44130.9	1.6				ug/L	27642	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	28483.4	4.0				ug/L	1647	Standard
P	31	20263.6	4.9				ug/L	15231	Standard
S	34	2148.5	1.6				ug/L	1845	Standard
Sr	88	1128.4	6.6				ug/L	70	Standard
C	12	163.3	25.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	36.0	33.6				mg/L	12	Standard
Ho-1	165	23.3	68.9				mg/L	12	Standard
Er	166	13.3	114.6				mg/L	20	Standard
I	127	19656.1	3.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		108.163	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.144	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606141602
Report Date/Time: Tuesday, June 28, 2016 21:26:25
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		90.979
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		89.007
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606142101
Sample Date/Time: Tuesday, June 28, 2016 21:27:19
Number of Replicates: 3
Autosampler Position: 327
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86072.2	1.7				ug/L	81760	Standard
	Be	9	31.7	65.7	**0.0204**	0.019	95.2	ug/L	17	Standard
	Al	27	37149346.8	0.4	**228.9677**	4.521	2.0	ug/L	1410	Standard
	Sc	45	40693.0	3.6				ug/L	39543	Standard
	Ti	47	211.7	4.5	**0.7055**	0.031	4.4	ug/L	39	Standard
	V	51	19470.4	0.9	**2.6593**	0.027	1.0	ug/L	1415	Standard
	Cr	52	9577.4	0.5	**0.3506**	0.020	5.8	ug/L	6946	Standard
	Cr	53	13504.4	2.9	**15.3965**	0.370	2.4	ug/L	328	Standard
	Mn	55	2482687.2	1.9	**391.6677**	8.445	2.2	ug/L	823	Standard
	Co	59	1010.4	4.1	**0.1158**	0.006	5.6	ug/L	252	Standard
	Ni	60	2375.2	1.3	**1.1855**	0.029	2.4	ug/L	192	Standard
	Cu	65	1289.7	5.2	**0.6034**	0.030	5.0	ug/L	141	Standard
	Zn	66	3124.3	3.5	**2.0882**	0.083	4.0	ug/L	491	Standard
>	**Ge**	72	503476.5	0.9				ug/L	515087	Standard
	As	75	9860.1	0.9	**9.0129**	0.111	1.2	ug/L	20	Standard
	Se	82	502.3	2.2	**4.3447**	0.140	3.2	ug/L	15	Standard
	Se-1	77	853.4	2.0	**10.8689**	0.134	1.2	ug/L	48	Standard
>	**Ga**	71	115.0	32.8				mg/L	90	Standard
	Rb	85	655.0	2.3				ug/L	23	Standard
	Y	89	413248.4	1.0				ug/L	448510	Standard
>	**Rh**	103	123.3	18.3				ug/L	15	Standard
	Mo	98	77.7	17.1	**0.0125**	0.006	46.7	ug/L	11	Standard
	Ag	107	81.3	4.3	**-0.0029**	0.001	26.3	ug/L	134	Standard
	Cd	111	15.2	10.0	**0.0049**	0.001	14.6	mg/L	7	Standard
	Cd	114	38.3	10.8	**0.0032**	0.001	30.4	ug/L	16	Standard
>	**In**	115	557968.8	1.5				ug/L	611117	Standard
	Sn	118	745.0	12.3	**0.0359**	0.015	41.7	ug/L	518	Standard
	Sb	123	991.0	1.5	**0.2172**	0.002	1.1	ug/L	127	Standard
	Ba	135	220850.7	1.1	**117.1448**	0.521	0.4	ug/L	45	Standard
	Ce	140	1566.7	2.4				ug/L	298	Standard
>	**Tb**	159	920767.6	2.2				ug/L	977942	Standard
	Ho	165	58.3	38.7				ug/L	12	Standard
	Tl	203	55.7	23.2	**0.0023**	0.002	74.7	ug/L	31	Standard
	Tl	205	45.0	38.5	**0.0043**	0.003	60.7	ug/L	20	Standard
	Pb	206	169197.9	0.8	**33.4745**	0.521	1.6	ug/L	286	Standard
	Pb	207	142071.0	0.5	**31.1598**	0.218	0.7	ug/L	257	Standard
	Pb	208	579992.3	2.5	**32.8772**	0.549	1.7	ug/L	956	Standard
	U	238	648.7	2.7	**0.0940**	0.004	4.0	ug/L	17	Standard
>	**Bi**	209	520066.6	1.1				ug/L	557080	Standard

Sample ID: L1606142101
Report Date/Time: Tuesday, June 28, 2016 21:29:36
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.0185**	17.344	173.1	mg/L	0	Standard
Mg	24	83474.3	1.3	**108.0258**	4.532	4.2	mg/L	32	Standard
K	39	53.3	28.6	**0.0335**	0.050	147.9	mg/L	30	Standard
Ca	43	435.0	11.3	**58.7409**	4.908	8.4	mg/L	48	Standard
Fe	54	1115.5	8.0	**0.8464**	0.111	13.2	mg/L	159	Standard
Fe	57	721.7	4.0	**1.7451**	0.132	7.5	mg/L	195	Standard
> **Sc-1**	45	40693.0	3.6				mg/L	39543	Standard
Cl	35	37196.3	1.1				ug/L	27642	Standard
Kr	83	3.0	66.7				ug/L	2	Standard
Br	81	117004.1	4.8				ug/L	1647	Standard
P	31	21849.1	2.8				ug/L	15231	Standard
S	34	2463.5	4.3				ug/L	1845	Standard
Sr	88	470.0	11.8				ug/L	70	Standard
C	12	400.0	4.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	47.0	20.8				mg/L	12	Standard
Ho-1	165	58.3	38.7				mg/L	12	Standard
Er	166	63.3	9.1				mg/L	20	Standard
I	127	290371.1	9.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		105.274	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.746	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606142101
Report Date/Time: Tuesday, June 28, 2016 21:29:36
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		91.303
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		93.356
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606142101
Report Date/Time: Tuesday, June 28, 2016 21:29:36
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 21:30:32
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	80064.5	3.2				ug/L	81760	Standard
	Be	9	49945.6	4.7	**51.3339**	1.054	2.1	ug/L	17	Standard
	Al	27	7446953.3	1.3	**49.3416**	1.021	2.1	ug/L	1410	Standard
	Sc	45	39994.4	3.1				ug/L	39543	Standard
	Ti	47	24563.6	1.8	**103.4611**	0.538	0.5	ug/L	39	Standard
	V	51	332459.7	1.4	**49.9760**	0.258	0.5	ug/L	1415	Standard
	Cr	52	346732.4	1.5	**51.2013**	0.219	0.4	ug/L	6946	Standard
	Cr	53	44092.4	3.1	**52.3486**	0.902	1.7	ug/L	328	Standard
	Mn	55	316773.7	1.0	**51.1225**	0.963	1.9	ug/L	823	Standard
	Co	59	338255.6	0.9	**50.4803**	0.452	0.9	ug/L	252	Standard
	Ni	60	91110.2	2.0	**50.6466**	0.186	0.4	ug/L	192	Standard
	Cu	65	91902.3	2.1	**50.6488**	0.160	0.3	ug/L	141	Standard
	Zn	66	53757.9	1.1	**50.6325**	0.663	1.3	ug/L	491	Standard
>	Ge	72	490984.9	1.8				ug/L	515087	Standard
	As	75	53512.8	1.1	**49.9767**	0.363	0.7	ug/L	20	Standard
	Se	82	5576.6	0.4	**50.7061**	0.891	1.8	ug/L	15	Standard
	Se-1	77	3693.5	2.8	**50.7124**	1.021	2.0	ug/L	48	Standard
>	Ga	71	63.3	29.9				mg/L	90	Standard
	Rb	85	831.7	5.0				ug/L	23	Standard
	Y	89	421530.5	1.2				ug/L	448510	Standard
>	Rh	103	28.3	44.4				ug/L	15	Standard
	Mo	98	254944.7	0.2	**97.5527**	0.365	0.4	ug/L	11	Standard
	Ag	107	312741.7	1.2	**48.1812**	0.687	1.4	ug/L	134	Standard
	Cd	111	100992.4	1.1	**48.8720**	0.473	1.0	mg/L	7	Standard
	Cd	114	246539.5	0.8	**49.0819**	0.317	0.6	ug/L	16	Standard
>	In	115	590755.4	0.3				ug/L	611117	Standard
	Sn	118	285297.8	0.9	**49.8102**	0.309	0.6	ug/L	518	Standard
	Sb	123	223087.7	1.5	**48.8998**	0.798	1.6	ug/L	127	Standard
	Ba	135	97286.0	1.5	**48.7240**	0.868	1.8	ug/L	45	Standard
	Ce	140	78.3	36.9				ug/L	298	Standard
>	Tb	159	940662.6	1.7				ug/L	977942	Standard
	Ho	165	20.0	25.0				ug/L	12	Standard
	Tl	203	402817.7	0.9	**49.5894**	0.553	1.1	ug/L	31	Standard
	Tl	205	361515.5	0.7	**51.2534**	0.688	1.3	ug/L	20	Standard
	Pb	206	256836.7	1.0	**49.2334**	0.527	1.1	ug/L	286	Standard
	Pb	207	233772.2	1.4	**49.6819**	0.304	0.6	ug/L	257	Standard
	Pb	208	939106.3	0.8	**51.5898**	0.228	0.4	ug/L	956	Standard
	U	238	351652.3	0.3	**50.9098**	0.427	0.8	ug/L	17	Standard
>	Bi	209	537054.5	1.2				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:32:49
Page 1

Approved: June 29, 2016

Na	23	5.0	173.2	**32.8471**	56.884	173.2	mg/L	0	Standard
Mg	24	4075.5	3.9	**5.3128**	0.340	6.4	mg/L	32	Standard
K	39	1373.4	2.2	**4.4475**	0.053	1.2	mg/L	30	Standard
Ca	43	71.7	22.4	**5.4062**	2.324	43.0	mg/L	48	Standard
Fe	54	5589.4	3.1	**4.8672**	0.205	4.2	mg/L	159	Standard
Fe	57	1635.1	4.9	**4.8331**	0.203	4.2	mg/L	195	Standard
Sc-1	45	39994.4	3.1				mg/L	39543	Standard
Cl	35	35754.8	4.0				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	3610.5	34.2				ug/L	1647	Standard
P	31	19324.0	4.2				ug/L	15231	Standard
S	34	2706.9	5.1				ug/L	1845	Standard
Sr	88	63.3	27.7				ug/L	70	Standard
C	12	143.3	42.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	5.9	100.8				mg/L	12	Standard
Ho-1	165	20.0	25.0				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	9711.8	44.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	102.668		
Al	27	98.683		
Sc	45			
Ti	47	103.461		
V	51	99.952		
Cr	52	102.403		
Cr	53			
Mn	55	102.245		
Co	59	100.961		
Ni	60	101.293		
Cu	65	101.298		
Zn	66	101.265		
Ge	72		95.321	
As	75	99.953		
Se	82	101.412		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:32:49
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	97.553		
Ag	107	96.362		
Cd	111	97.744		
Cd	114			
> In	115			96.668
Sn	118	99.620		
Sb	123	97.800		
Ba	135	97.448		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	99.179		
Tl	205			
Pb	206			
Pb	207			
Pb	208	103.180		
U	238	101.820		
> Bi	209			96.405
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 21:32:49
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 21:33:43
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77306.8	2.0				ug/L	81760	Standard
	Be	9	21.7	26.6	**0.0134**	0.006	48.3	ug/L	17	Standard
	Al	27	7303.5	27.4	**0.0355**	0.013	36.3	ug/L	1410	Standard
	Sc	45	38198.1	0.8				ug/L	39543	Standard
	Ti	47	25.7	25.3	**-0.0566**	0.027	47.6	ug/L	39	Standard
	V	51	1109.0	7.8	**-0.0386**	0.015	38.5	ug/L	1415	Standard
	Cr	52	6097.9	0.7	**-0.1350**	0.015	10.8	ug/L	6946	Standard
	Cr	53	763.4	6.1	**0.5733**	0.039	6.9	ug/L	328	Standard
	Mn	55	836.4	10.1	**-0.0128**	0.011	87.4	ug/L	823	Standard
	Co	59	230.3	21.5	**0.0031**	0.007	222.1	ug/L	252	Standard
	Ni	60	176.3	3.3	**-0.0064**	0.005	75.2	ug/L	192	Standard
	Cu	65	147.3	9.8	**-0.0093**	0.007	71.2	ug/L	141	Standard
	Zn	66	414.7	6.5	**-0.4297**	0.019	4.3	ug/L	491	Standard
>	**Ge**	72	489415.0	1.8				ug/L	515087	Standard
	As	75	-13.4	315.6	**0.0274**	0.040	146.9	ug/L	20	Standard
	Se	82	21.4	39.0	**0.0756**	0.076	100.0	ug/L	15	Standard
	Se-1	77	62.3	3.3	**0.1522**	0.035	23.1	ug/L	48	Standard
>	**Ga**	71	55.0	41.7				mg/L	90	Standard
	Rb	85	21.7	26.6				ug/L	23	Standard
	Y	89	425437.4	1.2				ug/L	448510	Standard
>	**Rh**	103	15.0	33.3				ug/L	15	Standard
	Mo	98	273.7	29.6	**0.0854**	0.031	36.3	ug/L	11	Standard
	Ag	107	107.7	30.0	**0.0003**	0.005	1550.1	ug/L	134	Standard
	Cd	111	12.2	52.2	**0.0030**	0.003	103.1	mg/L	7	Standard
	Cd	114	23.6	69.2	**-0.0002**	0.003	1385.1	ug/L	16	Standard
>	**In**	115	592746.5	0.3				ug/L	611117	Standard
	Sn	118	735.0	15.8	**0.0262**	0.020	76.7	ug/L	518	Standard
	Sb	123	418.9	32.2	**0.0787**	0.029	37.2	ug/L	127	Standard
	Ba	135	58.7	27.6	**0.0042**	0.008	193.4	ug/L	45	Standard
	Ce	140	16.7	34.6				ug/L	298	Standard
>	**Tb**	159	928666.7	1.6				ug/L	977942	Standard
	Ho	165	11.7	89.2				ug/L	12	Standard
	Tl	203	77.0	34.1	**0.0046**	0.003	68.4	ug/L	31	Standard
	Tl	205	76.7	35.9	**0.0085**	0.004	45.8	ug/L	20	Standard
	Pb	206	287.0	11.8	**-0.0088**	0.007	78.2	ug/L	286	Standard
	Pb	207	252.0	6.9	**-0.0047**	0.004	83.8	ug/L	257	Standard
	Pb	208	1038.7	6.8	**-0.0037**	0.004	102.1	ug/L	956	Standard
	U	238	52.3	19.6	**0.0046**	0.001	32.5	ug/L	17	Standard
>	**Bi**	209	540535.6	1.1				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:36:00
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**11.1254**	19.261	173.1	mg/L	0	Standard
Mg	24	41.7	25.0	**0.0002**	0.014	5858.2	mg/L	32	Standard
K	39	25.0	20.0	**-0.0539**	0.018	33.4	mg/L	30	Standard
Ca	43	40.0	25.0	**0.9508**	1.613	169.6	mg/L	48	Standard
Fe	54	117.8	25.5	**-0.0268**	0.028	102.8	mg/L	159	Standard
Fe	57	235.0	7.7	**0.1964**	0.057	28.8	mg/L	195	Standard
Sc-1	45	38198.1	0.8				mg/L	39543	Standard
Cl	35	35245.6	1.5				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	2320.2	7.7				ug/L	1647	Standard
P	31	18950.2	2.2				ug/L	15231	Standard
S	34	2475.2	4.2				ug/L	1845	Standard
Sr	88	68.3	27.7				ug/L	70	Standard
C	12	103.3	5.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	0	Standard
Dy	164	8.9	118.0				mg/L	12	Standard
Ho-1	165	11.7	89.2				mg/L	12	Standard
Er	166	23.3	49.5				mg/L	20	Standard
I	127	4469.0	13.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.016	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:36:00
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.994
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.030
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 21:36:00
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: PBW 82 WG574205-02
Sample Date/Time: Tuesday, June 28, 2016 21:36:55
Number of Replicates: 3
Autosampler Position: 328
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77750.7	1.6				ug/L	81760	Standard
	Be	9	23.3	24.7	**0.0149**	0.006	38.5	ug/L	17	Standard
	Al	27	11095.9	27.0	**0.0614**	0.021	33.8	ug/L	1410	Standard
	Sc	45	39309.3	3.2				ug/L	39543	Standard
	Ti	47	35.7	19.9	**-0.0160**	0.027	171.9	ug/L	39	Standard
	V	51	1269.4	3.8	**-0.0162**	0.012	71.1	ug/L	1415	Standard
	Cr	52	7452.2	2.1	**0.0582**	0.015	26.3	ug/L	6946	Standard
	Cr	53	805.0	4.3	**0.6140**	0.043	7.0	ug/L	328	Standard
	Mn	55	3268.0	4.0	**0.3772**	0.034	9.1	ug/L	823	Standard
	Co	59	296.7	4.7	**0.0126**	0.001	9.3	ug/L	252	Standard
	Ni	60	300.3	5.8	**0.0613**	0.013	21.5	ug/L	192	Standard
	Cu	65	252.7	4.2	**0.0478**	0.009	19.3	ug/L	141	Standard
	Zn	66	1868.1	1.9	**0.9480**	0.028	3.0	ug/L	491	Standard
>	**Ge**	72	494530.1	2.9				ug/L	515087	Standard
	As	75	-37.6	66.7	**0.0055**	0.023	429.2	ug/L	20	Standard
	Se	82	15.9	27.5	**0.0238**	0.038	158.5	ug/L	15	Standard
	Se-1	77	56.7	8.0	**0.0642**	0.051	79.2	ug/L	48	Standard
>	**Ga**	71	51.7	31.1				mg/L	90	Standard
	Rb	85	45.0	48.4				ug/L	23	Standard
	Y	89	413936.1	2.9				ug/L	448510	Standard
>	**Rh**	103	23.3	32.7				ug/L	15	Standard
	Mo	98	83.2	8.2	**0.0127**	0.003	23.0	ug/L	11	Standard
	Ag	107	95.3	10.0	**-0.0016**	0.002	102.9	ug/L	134	Standard
	Cd	111	39.5	14.4	**0.0162**	0.003	18.5	mg/L	7	Standard
	Cd	114	83.7	13.3	**0.0117**	0.002	21.2	ug/L	16	Standard
>	**In**	115	592344.0	1.8				ug/L	611117	Standard
	Sn	118	795.0	11.2	**0.0369**	0.017	47.0	ug/L	518	Standard
	Sb	123	176.9	20.5	**0.0258**	0.007	28.3	ug/L	127	Standard
	Ba	135	88.3	8.2	**0.0191**	0.004	22.5	ug/L	45	Standard
	Ce	140	48.3	15.8				ug/L	298	Standard
>	**Tb**	159	936174.2	1.5				ug/L	977942	Standard
	Ho	165	6.7	114.6				ug/L	12	Standard
	Tl	203	128.0	4.9	**0.0106**	0.001	7.1	ug/L	31	Standard
	Tl	205	116.7	44.6	**0.0139**	0.007	52.1	ug/L	20	Standard
	Pb	206	358.0	3.9	**0.0036**	0.003	73.1	ug/L	286	Standard
	Pb	207	341.0	6.5	**0.0129**	0.005	36.4	ug/L	257	Standard
	Pb	208	1292.4	1.6	**0.0090**	0.001	13.2	ug/L	956	Standard
	U	238	93.7	24.0	**0.0103**	0.003	31.1	ug/L	17	Standard
>	**Bi**	209	549438.7	0.1				ug/L	557080	Standard

Sample ID: PBW 82 WG574205-02
Report Date/Time: Tuesday, June 28, 2016 21:39:12
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.5912**	18.336	173.1	mg/L	0	Standard
Mg	24	55.0	47.2	**0.0170**	0.036	211.2	mg/L	32	Standard
K	39	38.3	52.7	**-0.0104**	0.070	678.8	mg/L	30	Standard
Ca	43	46.7	16.4	**1.7858**	1.139	63.8	mg/L	48	Standard
Fe	54	125.7	11.9	**-0.0224**	0.015	66.3	mg/L	159	Standard
Fe	57	226.7	12.1	**0.1455**	0.093	63.6	mg/L	195	Standard
> **Sc-1**	45	39309.3	3.2				mg/L	39543	Standard
Cl	35	34398.3	0.8				ug/L	27642	Standard
Kr	83	3.0	66.7				ug/L	2	Standard
Br	81	1940.1	5.9				ug/L	1647	Standard
P	31	18074.1	0.8				ug/L	15231	Standard
S	34	2498.5	3.0				ug/L	1845	Standard
Sr	88	61.7	46.1				ug/L	70	Standard
C	12	113.3	13.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	9.2	3.0				mg/L	12	Standard
Ho-1	165	6.7	114.6				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	3242.0	12.4				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		95.096	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		96.009	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBW 82 WG574205-02
Report Date/Time: Tuesday, June 28, 2016 21:39:12
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.928
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.628
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: LCSW 82 WG574205-03
Sample Date/Time: Tuesday, June 28, 2016 21:40:07
Number of Replicates: 3
Autosampler Position: 329
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77370.5	2.5				ug/L	81760	Standard
	Be	9	47069.7	3.6	**50.0685**	0.762	1.5	ug/L	17	Standard
	Al	27	8339.0	0.8	**0.0428**	0.001	2.3	ug/L	1410	Standard
	Sc	45	38442.1	2.3				ug/L	39543	Standard
	Ti	47	32.3	15.3	**-0.0274**	0.023	83.9	ug/L	39	Standard
	V	51	328373.9	1.3	**49.7728**	1.098	2.2	ug/L	1415	Standard
	Cr	52	340284.0	1.1	**50.6522**	0.528	1.0	ug/L	6946	Standard
	Cr	53	43268.4	1.2	**51.8066**	1.337	2.6	ug/L	328	Standard
	Mn	55	314954.5	0.2	**51.2439**	0.720	1.4	ug/L	823	Standard
	Co	59	331722.8	0.6	**49.9116**	0.628	1.3	ug/L	252	Standard
	Ni	60	91188.9	0.7	**51.1120**	0.365	0.7	ug/L	192	Standard
	Cu	65	91760.4	1.1	**50.9900**	0.227	0.4	ug/L	141	Standard
	Zn	66	54150.6	1.3	**51.4287**	0.106	0.2	ug/L	491	Standard
>	**Ge**	72	486985.3	1.4				ug/L	515087	Standard
	As	75	52577.5	0.5	**49.5070**	0.430	0.9	ug/L	20	Standard
	Se	82	5439.8	1.4	**49.8572**	0.058	0.1	ug/L	15	Standard
	Se-1	77	3674.8	1.0	**50.8806**	0.838	1.6	ug/L	48	Standard
>	**Ga**	71	70.0	21.4				mg/L	90	Standard
	Rb	85	35.0	37.8				ug/L	23	Standard
	Y	89	431830.9	0.7				ug/L	448510	Standard
>	**Rh**	103	38.3	15.1				ug/L	15	Standard
	Mo	98	69.3	10.5	**0.0074**	0.003	42.1	ug/L	11	Standard
	Ag	107	312423.9	0.2	**48.0443**	1.285	2.7	ug/L	134	Standard
	Cd	111	99250.2	0.2	**47.9394**	1.102	2.3	mg/L	7	Standard
	Cd	114	241436.3	0.5	**47.9735**	0.947	2.0	ug/L	16	Standard
>	**In**	115	592091.2	2.5				ug/L	611117	Standard
	Sn	118	838.4	9.3	**0.0444**	0.014	30.4	ug/L	518	Standard
	Sb	123	213126.5	0.4	**46.6279**	1.103	2.4	ug/L	127	Standard
	Ba	135	94941.2	1.6	**47.4531**	1.010	2.1	ug/L	45	Standard
	Ce	140	68.3	22.4				ug/L	298	Standard
>	**Tb**	159	920874.0	1.0				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	395784.4	0.5	**49.2144**	0.506	1.0	ug/L	31	Standard
	Tl	205	352148.8	0.9	**50.4246**	0.261	0.5	ug/L	20	Standard
	Pb	206	261292.0	1.2	**50.5931**	0.633	1.3	ug/L	286	Standard
	Pb	207	229140.8	1.1	**49.1895**	0.516	1.0	ug/L	257	Standard
	Pb	208	939702.2	0.3	**52.1447**	0.484	0.9	ug/L	956	Standard
	U	238	338970.6	0.7	**49.5659**	0.166	0.3	ug/L	17	Standard
>	**Bi**	209	531689.3	0.7				ug/L	557080	Standard

Sample ID: LCSW 82 WG574205-03
Report Date/Time: Tuesday, June 28, 2016 21:42:24
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	35.0	0.0	**-0.0092**	0.001	12.0	mg/L	32	Standard
K	39	35.0	42.9	**-0.0195**	0.052	268.3	mg/L	30	Standard
Ca	43	38.3	32.8	**0.6220**	1.885	303.1	mg/L	48	Standard
Fe	54	132.7	8.8	**-0.0133**	0.014	103.8	mg/L	159	Standard
Fe	57	226.7	10.4	**0.1628**	0.084	51.6	mg/L	195	Standard
> **Sc-1**	45	38442.1	2.3				mg/L	39543	Standard
Cl	35	35188.8	3.0				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1823.4	6.0				ug/L	1647	Standard
P	31	17611.9	0.6				ug/L	15231	Standard
S	34	2310.2	2.3				ug/L	1845	Standard
Sr	88	75.0	17.6				ug/L	70	Standard
C	12	96.7	46.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	5.9	194.5				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	16.7	34.6				mg/L	20	Standard
I	127	2716.9	12.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.631	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.544	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 82 WG574205-03
Report Date/Time: Tuesday, June 28, 2016 21:42:24
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.887
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.442
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 82 WG574205-03
Report Date/Time: Tuesday, June 28, 2016 21:42:24
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606134301 WG574205-01
Sample Date/Time: Tuesday, June 28, 2016 21:43:18
Number of Replicates: 3
Autosampler Position: 330
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74323.5	0.8				ug/L	81760	Standard
	Be	9	31.7	36.5	**0.0253**	0.013	49.8	ug/L	17	Standard
	Al	27	2998278.2	0.6	**21.3833**	0.050	0.2	ug/L	1410	Standard
	Sc	45	38510.6	3.3				ug/L	39543	Standard
	Ti	47	232.7	4.1	**0.8126**	0.032	3.9	ug/L	39	Standard
	V	51	1652.5	23.4	**0.0414**	0.055	131.6	ug/L	1415	Standard
	Cr	52	135321.3	0.7	**19.2638**	0.330	1.7	ug/L	6946	Standard
	Cr	53	19187.1	1.5	**22.5044**	0.281	1.2	ug/L	328	Standard
	Mn	55	3959.9	4.4	**0.4904**	0.040	8.1	ug/L	823	Standard
	Co	59	833.7	4.3	**0.0927**	0.008	8.3	ug/L	252	Standard
	Ni	60	44655.7	0.7	**24.6837**	0.632	2.6	ug/L	192	Standard
	Cu	65	31746.9	0.7	**17.3740**	0.276	1.6	ug/L	141	Standard
	Zn	66	57137.2	0.4	**53.6669**	1.121	2.1	ug/L	491	Standard
>	**Ge**	72	492860.3	1.9				ug/L	515087	Standard
	As	75	122.8	20.0	**0.1545**	0.023	14.7	ug/L	20	Standard
	Se	82	36.2	8.8	**0.2097**	0.035	16.8	ug/L	15	Standard
	Se-1	77	205.3	7.5	**2.1323**	0.253	11.9	ug/L	48	Standard
>	**Ga**	71	501.7	15.2				mg/L	90	Standard
	Rb	85	10964.0	0.9				ug/L	23	Standard
	Y	89	418533.1	1.8				ug/L	448510	Standard
>	**Rh**	103	56.7	31.0				ug/L	15	Standard
	Mo	98	2829.4	1.2	**1.1053**	0.009	0.8	ug/L	11	Standard
	Ag	107	25274.7	1.2	**4.0283**	0.033	0.8	ug/L	134	Standard
	Cd	111	119.1	9.1	**0.0569**	0.005	9.6	mg/L	7	Standard
	Cd	114	490.9	6.4	**0.0966**	0.006	6.7	ug/L	16	Standard
>	**In**	115	568910.2	0.4				ug/L	611117	Standard
	Sn	118	49674.7	2.3	**8.9227**	0.228	2.6	ug/L	518	Standard
	Sb	123	2415.2	17.6	**0.5368**	0.095	17.6	ug/L	127	Standard
	Ba	135	20410.1	0.9	**10.5946**	0.091	0.9	ug/L	45	Standard
	Ce	140	566.7	11.5				ug/L	298	Standard
>	**Tb**	159	895351.1	0.7				ug/L	977942	Standard
	Ho	165	20.0	25.0				ug/L	12	Standard
	Tl	203	216.3	14.7	**0.0231**	0.004	18.6	ug/L	31	Standard
	Tl	205	211.7	17.7	**0.0291**	0.005	18.1	ug/L	20	Standard
	Pb	206	14582.7	1.0	**2.8682**	0.053	1.8	ug/L	286	Standard
	Pb	207	11703.2	0.8	**2.5502**	0.014	0.5	ug/L	257	Standard
	Pb	208	49274.0	0.6	**2.7782**	0.042	1.5	ug/L	956	Standard
	U	238	116.0	23.9	**0.0146**	0.004	29.7	ug/L	17	Standard
>	**Bi**	209	512746.8	0.9				ug/L	557080	Standard

Sample ID: L1606134301 WG574205-01
Report Date/Time: Tuesday, June 28, 2016 21:45:35
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	23663.5	1.4	**32.3162**	1.385	4.3	mg/L	32	Standard
K	39	683.3	9.3	**2.2286**	0.197	8.8	mg/L	30	Standard
Ca	43	95.0	32.9	**9.3774**	4.330	46.2	mg/L	48	Standard
Fe	54	183.2	7.6	**0.0331**	0.009	27.1	mg/L	159	Standard
Fe	57	263.3	14.4	**0.2873**	0.119	41.4	mg/L	195	Standard
> **Sc-1**	45	38510.6	3.3				mg/L	39543	Standard
Cl	35	35170.8	1.9				ug/L	27642	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	3347.0	4.4				ug/L	1647	Standard
P	31	18034.1	2.9				ug/L	15231	Standard
S	34	2231.8	5.1				ug/L	1845	Standard
Sr	88	143.3	22.7				ug/L	70	Standard
C	12	123.3	47.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	12.4	90.1				mg/L	12	Standard
Ho-1	165	20.0	25.0				mg/L	12	Standard
Er	166	20.0	86.6				mg/L	20	Standard
I	127	33806.0	2.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		90.904	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.685	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606134301 WG574205-01
Report Date/Time: Tuesday, June 28, 2016 21:45:35
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		93.094
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		92.042
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606134301DP WG574205-04
Sample Date/Time: Tuesday, June 28, 2016 21:46:30
Number of Replicates: 3
Autosampler Position: 331
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75650.1	1.3				ug/L	81760	Standard
	Be	9	38.3	7.5	**0.0320**	0.004	11.5	ug/L	17	Standard
	Al	27	3034197.2	1.7	**21.2640**	0.568	2.7	ug/L	1410	Standard
	Sc	45	38358.5	1.6				ug/L	39543	Standard
	Ti	47	242.3	3.4	**0.8481**	0.030	3.6	ug/L	39	Standard
	V	51	2921.5	21.4	**0.2301**	0.091	39.7	ug/L	1415	Standard
	Cr	52	132102.7	0.7	**18.6755**	0.034	0.2	ug/L	6946	Standard
	Cr	53	19672.8	2.2	**22.9627**	0.596	2.6	ug/L	328	Standard
	Mn	55	13082.6	2.2	**1.9497**	0.036	1.8	ug/L	823	Standard
	Co	59	922.0	4.8	**0.1051**	0.006	5.6	ug/L	252	Standard
	Ni	60	44071.0	1.0	**24.2264**	0.187	0.8	ug/L	192	Standard
	Cu	65	31222.8	1.1	**16.9947**	0.096	0.6	ug/L	141	Standard
	Zn	66	55768.5	0.9	**52.0764**	0.197	0.4	ug/L	491	Standard
>	**Ge**	72	495382.5	0.5				ug/L	515087	Standard
	As	75	105.3	17.2	**0.1377**	0.017	12.5	ug/L	20	Standard
	Se	82	31.4	9.3	**0.1642**	0.025	15.2	ug/L	15	Standard
	Se-1	77	224.7	3.6	**2.3824**	0.126	5.3	ug/L	48	Standard
>	**Ga**	71	485.0	9.9				mg/L	90	Standard
	Rb	85	10640.4	0.2				ug/L	23	Standard
	Y	89	407749.6	1.6				ug/L	448510	Standard
>	**Rh**	103	56.7	27.0				ug/L	15	Standard
	Mo	98	2759.9	2.4	**1.0902**	0.013	1.1	ug/L	11	Standard
	Ag	107	24718.5	0.6	**3.9851**	0.067	1.7	ug/L	134	Standard
	Cd	111	119.1	7.2	**0.0577**	0.005	8.8	mg/L	7	Standard
	Cd	114	389.6	30.3	**0.0764**	0.024	30.9	ug/L	16	Standard
>	**In**	115	562489.2	1.2				ug/L	611117	Standard
	Sn	118	48104.7	5.2	**8.7400**	0.532	6.1	ug/L	518	Standard
	Sb	123	1085.9	11.8	**0.2370**	0.026	11.2	ug/L	127	Standard
	Ba	135	21279.9	0.9	**11.1745**	0.156	1.4	ug/L	45	Standard
	Ce	140	1446.7	1.7				ug/L	298	Standard
>	**Tb**	159	899809.7	0.9				ug/L	977942	Standard
	Ho	165	38.3	37.7				ug/L	12	Standard
	Tl	203	211.3	8.9	**0.0223**	0.003	11.4	ug/L	31	Standard
	Tl	205	171.7	14.4	**0.0230**	0.004	16.6	ug/L	20	Standard
	Pb	206	14329.4	1.7	**2.7934**	0.069	2.5	ug/L	286	Standard
	Pb	207	11471.7	1.5	**2.4775**	0.050	2.0	ug/L	257	Standard
	Pb	208	48196.5	1.3	**2.6929**	0.035	1.3	ug/L	956	Standard
	U	238	162.0	9.3	**0.0214**	0.002	11.2	ug/L	17	Standard
>	**Bi**	209	517046.0	0.7				ug/L	557080	Standard

Sample ID: L1606134301DP WG574205-04
Report Date/Time: Tuesday, June 28, 2016 21:48:47
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	23364.7	1.2	**32.0144**	0.886	2.8	mg/L	32	Standard
K	39	620.0	6.9	**2.0197**	0.176	8.7	mg/L	30	Standard
Ca	43	98.3	39.5	**10.0220**	6.022	60.1	mg/L	48	Standard
Fe	54	175.4	14.5	**0.0269**	0.026	96.7	mg/L	159	Standard
Fe	57	298.3	7.6	**0.4146**	0.094	22.7	mg/L	195	Standard
> **Sc-1**	45	38358.5	1.6				mg/L	39543	Standard
Cl	35	35843.0	1.2				ug/L	27642	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	4177.2	4.7				ug/L	1647	Standard
P	31	18070.8	1.2				ug/L	15231	Standard
S	34	2273.5	5.8				ug/L	1845	Standard
Sr	88	125.0	20.0				ug/L	70	Standard
C	12	143.3	21.3				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	16.7	91.7				mg/L	0	Standard
Dy	164	45.4	50.3				mg/L	12	Standard
Ho-1	165	38.3	37.7				mg/L	12	Standard
Er	166	26.7	21.7				mg/L	20	Standard
I	127	32309.4	0.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		92.527	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.175	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606134301DP WG574205-04
Report Date/Time: Tuesday, June 28, 2016 21:48:47
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		92.043
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		92.814
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606134301S WG574205-05
Sample Date/Time: Tuesday, June 28, 2016 21:49:41
Number of Replicates: 3
Autosampler Position: 332
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	76159.4	2.8				ug/L	81760	Standard
	Be	9	48450.7	1.7	**52.3785**	0.949	1.8	ug/L	17	Standard
	Al	27	2895043.9	1.0	**20.1618**	0.737	3.7	ug/L	1410	Standard
	Sc	45	38742.8	1.9				ug/L	39543	Standard
	Ti	47	240.0	7.2	**0.8720**	0.082	9.4	ug/L	39	Standard
	V	51	323841.2	0.8	**49.8512**	0.902	1.8	ug/L	1415	Standard
	Cr	52	453207.8	1.4	**68.8842**	0.591	0.9	ug/L	6946	Standard
	Cr	53	60472.1	1.8	**73.6607**	0.780	1.1	ug/L	328	Standard
	Mn	55	309167.0	1.2	**51.0846**	0.705	1.4	ug/L	823	Standard
	Co	59	322466.2	0.8	**49.2767**	0.754	1.5	ug/L	252	Standard
	Ni	60	129806.7	1.0	**73.9391**	0.553	0.7	ug/L	192	Standard
	Cu	65	118360.2	1.2	**66.8253**	0.317	0.5	ug/L	141	Standard
	Zn	66	105137.5	0.6	**102.2235**	1.030	1.0	ug/L	491	Standard
>	Ge	72	479488.8	1.1				ug/L	515087	Standard
	As	75	53941.4	1.7	**51.5891**	1.400	2.7	ug/L	20	Standard
	Se	82	5350.1	0.9	**49.8082**	0.982	2.0	ug/L	15	Standard
	Se-1	77	3727.1	2.1	**52.4361**	1.513	2.9	ug/L	48	Standard
>	Ga	71	476.7	2.2				mg/L	90	Standard
	Rb	85	10533.7	3.9				ug/L	23	Standard
	Y	89	417307.7	1.3				ug/L	448510	Standard
>	Rh	103	73.3	25.8				ug/L	15	Standard
	Mo	98	2762.6	0.6	**1.0591**	0.012	1.2	ug/L	11	Standard
	Ag	107	321242.3	0.9	**50.4672**	0.741	1.5	ug/L	134	Standard
	Cd	111	96845.5	0.4	**47.7898**	0.498	1.0	mg/L	7	Standard
	Cd	114	233334.7	1.1	**47.3688**	0.646	1.4	ug/L	16	Standard
>	In	115	579360.0	0.8				ug/L	611117	Standard
	Sn	118	55146.4	13.8	**9.7304**	1.292	13.3	ug/L	518	Standard
	Sb	123	221114.9	0.1	**49.4220**	0.385	0.8	ug/L	127	Standard
	Ba	135	112895.7	0.5	**57.6594**	0.626	1.1	ug/L	45	Standard
	Ce	140	313.3	11.5				ug/L	298	Standard
>	Tb	159	912553.5	0.8				ug/L	977942	Standard
	Ho	165	15.0	66.7				ug/L	12	Standard
	Tl	203	384933.8	0.5	**49.0255**	0.608	1.2	ug/L	31	Standard
	Tl	205	344367.9	0.4	**50.5097**	0.783	1.5	ug/L	20	Standard
	Pb	206	266968.5	0.3	**52.9500**	0.690	1.3	ug/L	286	Standard
	Pb	207	232893.3	0.6	**51.2096**	0.463	0.9	ug/L	257	Standard
	Pb	208	948808.9	1.4	**53.9303**	1.097	2.0	ug/L	956	Standard
	U	238	335133.4	0.9	**50.1950**	0.747	1.5	ug/L	17	Standard
>	Bi	209	519140.8	1.4				ug/L	557080	Standard

Sample ID: L1606134301S WG574205-05
Report Date/Time: Tuesday, June 28, 2016 21:51:58
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	23122.6	1.5	**31.3670**	0.821	2.6	mg/L	32	Standard
K	39	550.0	13.4	**1.7528**	0.224	12.8	mg/L	30	Standard
Ca	43	93.3	3.1	**9.1209**	0.222	2.4	mg/L	48	Standard
Fe	54	147.8	38.5	**0.0002**	0.055	34709.4	mg/L	159	Standard
Fe	57	261.7	11.7	**0.2782**	0.120	43.1	mg/L	195	Standard
> **Sc-1**	45	38742.8	1.9				mg/L	39543	Standard
Cl	35	35887.1	0.6				ug/L	27642	Standard
Kr	83	2.3	89.2				ug/L	2	Standard
Br	81	3013.6	4.3				ug/L	1647	Standard
P	31	18511.3	3.1				ug/L	15231	Standard
S	34	2173.5	7.9				ug/L	1845	Standard
Sr	88	115.0	4.3				ug/L	70	Standard
C	12	143.3	35.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	12.7	50.0				mg/L	12	Standard
Ho-1	165	15.0	66.7				mg/L	12	Standard
Er	166	13.3	114.6				mg/L	20	Standard
I	127	29318.3	1.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.150	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.089	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606134301S WG574205-05
Report Date/Time: Tuesday, June 28, 2016 21:51:58
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.804
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		93.190
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Zn 66 Upper, S, EEE	Zn	66	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606134301SD WG574205-06
Sample Date/Time: Tuesday, June 28, 2016 21:52:53
Number of Replicates: 3
Autosampler Position: 333
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74233.3	4.4				ug/L	81760	Standard
	Be	9	46944.3	0.7	**52.1161**	2.208	4.2	ug/L	17	Standard
	Al	27	2867977.1	1.8	**20.5130**	1.211	5.9	ug/L	1410	Standard
	Sc	45	37107.1	4.7				ug/L	39543	Standard
	Ti	47	208.3	2.4	**0.7439**	0.023	3.1	ug/L	39	Standard
	V	51	314357.5	0.7	**48.8632**	0.269	0.5	ug/L	1415	Standard
	Cr	52	443013.9	1.2	**67.9920**	0.699	1.0	ug/L	6946	Standard
	Cr	53	57726.4	1.5	**71.0049**	0.771	1.1	ug/L	328	Standard
	Mn	55	299003.6	1.5	**49.8899**	0.541	1.1	ug/L	823	Standard
	Co	59	310607.5	1.9	**47.9296**	0.658	1.4	ug/L	252	Standard
	Ni	60	125816.0	1.4	**72.3742**	0.740	1.0	ug/L	192	Standard
	Cu	65	116930.9	2.0	**66.6728**	1.005	1.5	ug/L	141	Standard
	Zn	66	105389.9	0.9	**103.4925**	0.359	0.3	ug/L	491	Standard
>	**Ge**	72	474757.8	0.5				ug/L	515087	Standard
	As	75	52320.0	1.0	**50.5276**	0.214	0.4	ug/L	20	Standard
	Se	82	5199.8	1.3	**48.8839**	0.776	1.6	ug/L	15	Standard
	Se-1	77	3645.4	2.7	**51.7794**	1.336	2.6	ug/L	48	Standard
>	**Ga**	71	471.7	1.2				mg/L	90	Standard
	Rb	85	10923.9	0.7				ug/L	23	Standard
	Y	89	414912.1	1.8				ug/L	448510	Standard
>	**Rh**	103	51.7	22.3				ug/L	15	Standard
	Mo	98	2810.6	3.1	**1.1098**	0.050	4.5	ug/L	11	Standard
	Ag	107	309794.3	1.1	**50.0964**	1.925	3.8	ug/L	134	Standard
	Cd	111	92939.1	0.1	**47.2004**	1.405	3.0	mg/L	7	Standard
	Cd	114	227712.2	0.8	**47.5835**	1.757	3.7	ug/L	16	Standard
>	**In**	115	563239.5	3.0				ug/L	611117	Standard
	Sn	118	48800.3	6.0	**8.8657**	0.749	8.5	ug/L	518	Standard
	Sb	123	213681.9	0.5	**49.1574**	1.598	3.3	ug/L	127	Standard
	Ba	135	108173.8	0.8	**56.8630**	1.899	3.3	ug/L	45	Standard
	Ce	140	258.3	11.8				ug/L	298	Standard
>	**Tb**	159	874543.1	3.3				ug/L	977942	Standard
	Ho	165	16.7	62.4				ug/L	12	Standard
	Tl	203	370949.6	0.9	**49.2226**	1.185	2.4	ug/L	31	Standard
	Tl	205	332170.0	0.6	**50.7579**	1.084	2.1	ug/L	20	Standard
	Pb	206	260124.0	0.8	**53.7518**	1.191	2.2	ug/L	286	Standard
	Pb	207	226333.8	1.1	**51.8538**	1.336	2.6	ug/L	257	Standard
	Pb	208	924176.8	1.2	**54.7310**	1.509	2.8	ug/L	956	Standard
	U	238	327055.2	1.2	**51.0349**	1.235	2.4	ug/L	17	Standard
>	**Bi**	209	498348.0	1.6				ug/L	557080	Standard

Sample ID: L1606134301SD WG574205-06
Report Date/Time: Tuesday, June 28, 2016 21:55:09
Page 1

Approved: June 29, 2016

Na	23	5.0	173.2	**33.3751**	57.799	173.2	mg/L	0	Standard
Mg	24	22616.9	1.7	**32.0867**	1.983	6.2	mg/L	32	Standard
K	39	585.0	8.2	**1.9630**	0.088	4.5	mg/L	30	Standard
Ca	43	106.7	21.1	**11.9418**	3.627	30.4	mg/L	48	Standard
Fe	54	143.6	9.6	**0.0016**	0.013	810.1	mg/L	159	Standard
Fe	57	260.0	25.0	**0.3188**	0.279	87.4	mg/L	195	Standard
> **Sc-1**	45	37107.1	4.7				mg/L	39543	Standard
Cl	35	35012.4	2.2				ug/L	27642	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	3463.7	1.0				ug/L	1647	Standard
P	31	17954.0	1.9				ug/L	15231	Standard
S	34	2235.2	3.1				ug/L	1845	Standard
Sr	88	115.0	27.2				ug/L	70	Standard
C	12	173.3	16.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	0	Standard
Dy	164	16.2	97.3				mg/L	12	Standard
Ho-1	165	16.7	62.4				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	29468.6	2.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		90.794	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.170	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606134301SD WG574205-06
Report Date/Time: Tuesday, June 28, 2016 21:55:09
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		92.166
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		89.457
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606134301SD WG574205-06
Report Date/Time: Tuesday, June 28, 2016 21:55:09
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606073201
Sample Date/Time: Tuesday, June 28, 2016 21:56:04
Number of Replicates: 3
Autosampler Position: 334
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79995.7	1.9				ug/L	81760	Standard
	Be	9	111.7	43.0	**0.1055**	0.050	47.4	ug/L	17	Standard
	Al	27	26418031.8	2.7	**175.1861**	5.381	3.1	ug/L	1410	Standard
	Sc	45	39758.8	3.3				ug/L	39543	Standard
	Ti	47	586.7	3.3	**2.3079**	0.104	4.5	ug/L	39	Standard
	V	51	2345.3	13.2	**0.1468**	0.044	29.9	ug/L	1415	Standard
	Cr	52	26092.8	3.1	**2.8709**	0.089	3.1	ug/L	6946	Standard
	Cr	53	22536.8	6.4	**26.5559**	1.479	5.6	ug/L	328	Standard
	Mn	55	128770.6	1.1	**20.6681**	0.068	0.3	ug/L	823	Standard
	Co	59	4370.6	1.2	**0.6206**	0.009	1.4	ug/L	252	Standard
	Ni	60	4217.9	1.8	**2.2421**	0.029	1.3	ug/L	192	Standard
	Cu	65	1551.7	2.2	**0.7650**	0.010	1.3	ug/L	141	Standard
	Zn	66	53557.8	1.5	**50.3807**	0.333	0.7	ug/L	491	Standard
>	**Ge**	72	491490.5	1.0				ug/L	515087	Standard
	As	75	918.5	14.0	**0.8957**	0.112	12.5	ug/L	20	Standard
	Se	82	316.6	5.4	**2.7619**	0.129	4.7	ug/L	15	Standard
	Se-1	77	1234.4	4.8	**16.4500**	0.704	4.3	ug/L	48	Standard
>	**Ga**	71	80.0	6.3				mg/L	90	Standard
	Rb	85	7720.3	1.4				ug/L	23	Standard
	Y	89	403861.7	0.4				ug/L	448510	Standard
>	**Rh**	103	80.0	50.0				ug/L	15	Standard
	Mo	98	355.0	5.6	**0.1283**	0.008	5.9	ug/L	11	Standard
	Ag	107	141.3	9.2	**0.0074**	0.002	31.2	ug/L	134	Standard
	Cd	111	202.7	10.3	**0.1036**	0.012	11.2	mg/L	7	Standard
	Cd	114	507.6	13.6	**0.1047**	0.015	13.9	ug/L	16	Standard
>	**In**	115	544497.8	0.6				ug/L	611117	Standard
	Sn	118	1023.4	24.9	**0.0925**	0.050	53.7	ug/L	518	Standard
	Sb	123	669.1	12.6	**0.1463**	0.019	13.2	ug/L	127	Standard
	Ba	135	206763.5	2.5	**112.3722**	2.187	1.9	ug/L	45	Standard
	Ce	140	411.7	6.1				ug/L	298	Standard
>	**Tb**	159	889575.8	1.3				ug/L	977942	Standard
	Ho	165	35.0	37.8				ug/L	12	Standard
	Tl	203	749.0	2.3	**0.0953**	0.003	3.6	ug/L	31	Standard
	Tl	205	685.0	13.9	**0.1029**	0.013	12.2	ug/L	20	Standard
	Pb	206	776.0	1.1	**0.0981**	0.006	5.8	ug/L	286	Standard
	Pb	207	670.3	1.0	**0.0968**	0.005	4.8	ug/L	257	Standard
	Pb	208	2692.4	0.9	**0.1003**	0.005	4.6	ug/L	956	Standard
	U	238	1446.7	3.3	**0.2244**	0.006	2.7	ug/L	17	Standard
>	**Bi**	209	494871.3	2.5				ug/L	557080	Standard

Sample ID: L1606073201
Report Date/Time: Tuesday, June 28, 2016 21:58:21
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.6945**	18.515	173.1	mg/L	0	Standard
Mg	24	122893.2	2.3	**162.6797**	1.668	1.0	mg/L	32	Standard
K	39	163.3	4.7	**0.4080**	0.035	8.6	mg/L	30	Standard
Ca	43	160.0	24.8	**18.8525**	6.186	32.8	mg/L	48	Standard
Fe	54	210.9	15.7	**0.0529**	0.031	58.0	mg/L	159	Standard
Fe	57	311.7	8.2	**0.4207**	0.052	12.3	mg/L	195	Standard
> **Sc-1**	45	39758.8	3.3				mg/L	39543	Standard
Cl	35	36983.1	0.7				ug/L	27642	Standard
Kr	83	2.3	49.5				ug/L	2	Standard
Br	81	75233.3	4.9				ug/L	1647	Standard
P	31	24466.4	2.9				ug/L	15231	Standard
S	34	2258.5	3.0				ug/L	1845	Standard
Sr	88	378.3	24.5				ug/L	70	Standard
C	12	350.0	10.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	120.9	45.9				mg/L	12	Standard
Ho-1	165	35.0	37.8				mg/L	12	Standard
Er	166	50.0	20.0				mg/L	20	Standard
I	127	539973.2	10.6				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		97.842	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		95.419	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606073201
Report Date/Time: Tuesday, June 28, 2016 21:58:21
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		89.099
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		88.833
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Ba 135 Upper, S, EEE	Ba	135	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606073201PS WG574571-01
Sample Date/Time: Tuesday, June 28, 2016 21:59:15
Number of Replicates: 3
Autosampler Position: 335
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83254.8	1.5				ug/L	81760	Standard
	Be	9	50562.6	1.1	**49.9976**	0.969	1.9	ug/L	17	Standard
	Al	27	27107676.6	2.1	**172.7073**	3.970	2.3	ug/L	1410	Standard
	Sc	45	41809.3	4.0				ug/L	39543	Standard
	Ti	47	632.7	1.1	**2.4753**	0.055	2.2	ug/L	39	Standard
	V	51	341856.0	1.3	**50.8320**	0.708	1.4	ug/L	1415	Standard
	Cr	52	363663.1	1.5	**53.1642**	1.402	2.6	ug/L	6946	Standard
	Cr	53	65238.5	1.6	**76.7756**	0.750	1.0	ug/L	328	Standard
	Mn	55	438903.0	0.5	**70.1258**	2.130	3.0	ug/L	823	Standard
	Co	59	331676.0	1.2	**48.9632**	1.267	2.6	ug/L	252	Standard
	Ni	60	91787.9	1.2	**50.4717**	0.832	1.6	ug/L	192	Standard
	Cu	65	90254.4	2.3	**49.1926**	0.475	1.0	ug/L	141	Standard
	Zn	66	107316.4	1.1	**100.7774**	1.488	1.5	ug/L	491	Standard
>	**Ge**	72	496483.6	2.5				ug/L	515087	Standard
	As	75	55410.9	1.5	**51.1805**	0.552	1.1	ug/L	20	Standard
	Se	82	5728.8	0.9	**51.5315**	1.616	3.1	ug/L	15	Standard
	Se-1	77	4873.5	1.7	**66.4041**	0.559	0.8	ug/L	48	Standard
>	**Ga**	71	86.7	34.8				mg/L	90	Standard
	Rb	85	8242.2	2.9				ug/L	23	Standard
	Y	89	419418.3	1.5				ug/L	448510	Standard
>	**Rh**	103	96.7	18.2				ug/L	15	Standard
	Mo	98	373.5	3.6	**0.1304**	0.006	4.5	ug/L	11	Standard
	Ag	107	276030.5	2.3	**44.4706**	1.418	3.2	ug/L	134	Standard
	Cd	111	97190.1	1.0	**49.1769**	0.174	0.4	mg/L	7	Standard
	Cd	114	234963.3	0.8	**48.9150**	0.797	1.6	ug/L	16	Standard
>	**In**	115	564991.8	0.9				ug/L	611117	Standard
	Sn	118	1183.4	5.3	**0.1145**	0.011	9.2	ug/L	518	Standard
	Sb	123	221298.1	1.9	**50.7268**	1.390	2.7	ug/L	127	Standard
	Ba	135	302634.3	0.9	**158.5372**	1.746	1.1	ug/L	45	Standard
	Ce	140	460.0	11.5				ug/L	298	Standard
>	**Tb**	159	913205.8	1.2				ug/L	977942	Standard
	Ho	165	30.0	16.7				ug/L	12	Standard
	Tl	203	392737.2	0.1	**50.5926**	0.297	0.6	ug/L	31	Standard
	Tl	205	346685.2	1.3	**51.4296**	0.576	1.1	ug/L	20	Standard
	Pb	206	257242.2	1.9	**51.6050**	1.169	2.3	ug/L	286	Standard
	Pb	207	223447.5	1.6	**49.6960**	0.910	1.8	ug/L	257	Standard
	Pb	208	904356.3	0.4	**51.9882**	0.250	0.5	ug/L	956	Standard
	U	238	346838.0	0.8	**52.5438**	0.601	1.1	ug/L	17	Standard
>	**Bi**	209	513215.0	0.6				ug/L	557080	Standard

Sample ID: L1606073201PS WG574571-01
Report Date/Time: Tuesday, June 28, 2016 22:01:32
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**9.7143**	16.817	173.1	mg/L	0	Standard
Mg	24	126816.9	1.1	**159.8003**	7.491	4.7	mg/L	32	Standard
K	39	205.0	11.2	**0.5135**	0.066	12.9	mg/L	30	Standard
Ca	43	170.0	5.9	**19.0779**	1.160	6.1	mg/L	48	Standard
Fe	54	195.8	12.4	**0.0303**	0.016	52.3	mg/L	159	Standard
Fe	57	350.0	26.8	**0.4863**	0.252	51.8	mg/L	195	Standard
> **Sc-1**	45	41809.3	4.0				mg/L	39543	Standard
Cl	35	38499.5	1.5				ug/L	27642	Standard
Kr	83	2.7	78.1				ug/L	2	Standard
Br	81	75812.6	2.8				ug/L	1647	Standard
P	31	25067.4	2.7				ug/L	15231	Standard
S	34	2465.2	4.5				ug/L	1845	Standard
Sr	88	373.3	2.8				ug/L	70	Standard
C	12	380.0	17.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	29.2	88.7				mg/L	12	Standard
Ho-1	165	30.0	16.7				mg/L	12	Standard
Er	166	16.7	69.3				mg/L	20	Standard
I	127	574284.0	7.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.828	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.388	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606073201PS WG574571-01
Report Date/Time: Tuesday, June 28, 2016 22:01:32
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	92.452
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	92.126
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606073201PS WG574571-01
Report Date/Time: Tuesday, June 28, 2016 22:01:32
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606073202SDL WG574571-02
Sample Date/Time: Tuesday, June 28, 2016 22:02:27
Number of Replicates: 3
Autosampler Position: 336
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80721.2	1.7				ug/L	81760	Standard
	Be	9	38.3	19.9	**0.0293**	0.007	25.3	ug/L	17	Standard
	Al	27	5438081.5	1.8	**35.7222**	0.611	1.7	ug/L	1410	Standard
	Sc	45	40838.3	2.8				ug/L	39543	Standard
	Ti	47	147.3	1.4	**0.4611**	0.014	3.1	ug/L	39	Standard
	V	51	1818.8	6.0	**0.0702**	0.022	30.8	ug/L	1415	Standard
	Cr	52	11457.0	0.2	**0.6827**	0.037	5.5	ug/L	6946	Standard
	Cr	53	6112.9	0.7	**7.0202**	0.190	2.7	ug/L	328	Standard
	Mn	55	27400.4	0.7	**4.3193**	0.101	2.3	ug/L	823	Standard
	Co	59	1102.0	5.1	**0.1343**	0.005	3.9	ug/L	252	Standard
	Ni	60	1029.7	4.0	**0.4727**	0.014	2.9	ug/L	192	Standard
	Cu	65	552.3	3.6	**0.2164**	0.013	6.0	ug/L	141	Standard
	Zn	66	17277.2	1.2	**15.8333**	0.273	1.7	ug/L	491	Standard
>	**Ge**	72	487411.0	2.0				ug/L	515087	Standard
	As	75	241.2	19.0	**0.2666**	0.040	15.0	ug/L	20	Standard
	Se	82	96.0	8.9	**0.7616**	0.066	8.7	ug/L	15	Standard
	Se-1	77	318.7	3.3	**3.7527**	0.181	4.8	ug/L	48	Standard
>	**Ga**	71	61.7	24.8				mg/L	90	Standard
	Rb	85	1698.4	2.7				ug/L	23	Standard
	Y	89	428587.6	1.1				ug/L	448510	Standard
>	**Rh**	103	33.3	37.7				ug/L	15	Standard
	Mo	98	99.2	5.5	**0.0185**	0.002	11.0	ug/L	11	Standard
	Ag	107	280.0	31.3	**0.0265**	0.013	49.6	ug/L	134	Standard
	Cd	111	57.2	23.8	**0.0245**	0.006	26.0	mg/L	7	Standard
	Cd	114	159.7	10.5	**0.0266**	0.003	11.9	ug/L	16	Standard
>	**In**	115	596430.5	0.6				ug/L	611117	Standard
	Sn	118	631.7	19.2	**0.0076**	0.022	282.9	ug/L	518	Standard
	Sb	123	2037.7	21.9	**0.4301**	0.100	23.2	ug/L	127	Standard
	Ba	135	42595.5	1.1	**21.1150**	0.168	0.8	ug/L	45	Standard
	Ce	140	145.0	24.9				ug/L	298	Standard
>	**Tb**	159	930009.8	1.5				ug/L	977942	Standard
	Ho	165	10.0					ug/L	12	Standard
	Tl	203	287.0	3.9	**0.0311**	0.001	3.3	ug/L	31	Standard
	Tl	205	238.3	17.8	**0.0321**	0.006	20.0	ug/L	20	Standard
	Pb	206	467.7	4.1	**0.0276**	0.003	10.6	ug/L	286	Standard
	Pb	207	403.3	2.0	**0.0292**	0.001	2.3	ug/L	257	Standard
	Pb	208	1685.7	2.5	**0.0338**	0.002	5.7	ug/L	956	Standard
	U	238	344.0	11.7	**0.0476**	0.005	11.0	ug/L	17	Standard
>	**Bi**	209	528241.9	1.3				ug/L	557080	Standard

Sample ID: L1606073202SDL WG574571-02
Report Date/Time: Tuesday, June 28, 2016 22:04:44
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	25095.8	1.4	**32.3042**	0.912	2.8	mg/L	32	Standard
K	39	53.3	28.6	**0.0336**	0.053	158.1	mg/L	30	Standard
Ca	43	51.7	14.8	**2.2359**	0.907	40.5	mg/L	48	Standard
Fe	54	140.6	10.4	**-0.0136**	0.016	114.2	mg/L	159	Standard
Fe	57	253.3	13.1	**0.2033**	0.104	51.3	mg/L	195	Standard
> **Sc-1**	45	40838.3	2.8				mg/L	39543	Standard
Cl	35	38490.2	0.9				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	17116.3	3.1				ug/L	1647	Standard
P	31	18993.6	2.2				ug/L	15231	Standard
S	34	2316.8	7.4				ug/L	1845	Standard
Sr	88	121.7	17.1				ug/L	70	Standard
C	12	150.0	13.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	5.2	126.3				mg/L	12	Standard
Ho-1	165	10.0					mg/L	12	Standard
Er	166	30.0	57.7				mg/L	20	Standard
I	127	133767.9	0.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		98.729	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.627	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606073202SDL WG574571-02
Report Date/Time: Tuesday, June 28, 2016 22:04:44
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.597
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.823
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606073202SDL WG574571-02
Sample Date/Time: Tuesday, June 28, 2016 22:05:38
Number of Replicates: 3
Autosampler Position: 337
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76053.8	2.7				ug/L	81760	Standard
	Be	9	25.0	34.6	**0.0174**	0.009	54.5	ug/L	17	Standard
	Al	27	1128796.7	0.4	**7.8615**	0.198	2.5	ug/L	1410	Standard
	Sc	45	38686.0	0.8				ug/L	39543	Standard
	Ti	47	55.3	34.5	**0.0617**	0.078	127.2	ug/L	39	Standard
	V	51	1638.9	5.5	**0.0337**	0.013	38.4	ug/L	1415	Standard
	Cr	52	8717.5	0.5	**0.2193**	0.011	4.9	ug/L	6946	Standard
	Cr	53	2070.1	6.6	**2.0621**	0.160	7.7	ug/L	328	Standard
	Mn	55	6629.8	1.0	**0.8938**	0.008	0.9	ug/L	823	Standard
	Co	59	385.0	3.7	**0.0245**	0.002	8.1	ug/L	252	Standard
	Ni	60	384.7	1.7	**0.1032**	0.003	2.7	ug/L	192	Standard
	Cu	65	339.7	2.9	**0.0913**	0.005	5.2	ug/L	141	Standard
	Zn	66	9883.2	1.3	**8.3627**	0.143	1.7	ug/L	491	Standard
>	**Ge**	72	505388.6	0.3				ug/L	515087	Standard
	As	75	-13.0	254.9	**0.0286**	0.030	104.7	ug/L	20	Standard
	Se	82	27.4	29.7	**0.1233**	0.072	58.7	ug/L	15	Standard
	Se-1	77	108.0	8.5	**0.7425**	0.129	17.3	ug/L	48	Standard
>	**Ga**	71	48.3	46.6				mg/L	90	Standard
	Rb	85	370.0	8.2				ug/L	23	Standard
	Y	89	428119.2	2.9				ug/L	448510	Standard
>	**Rh**	103	23.3	44.6				ug/L	15	Standard
	Mo	98	45.2	4.1	**-0.0015**	0.001	60.8	ug/L	11	Standard
	Ag	107	113.3	16.4	**0.0016**	0.003	172.3	ug/L	134	Standard
	Cd	111	14.3	24.6	**0.0041**	0.002	39.1	mg/L	7	Standard
	Cd	114	49.9	58.5	**0.0052**	0.006	113.8	ug/L	16	Standard
>	**In**	115	579695.7	1.7				ug/L	611117	Standard
	Sn	118	626.7	2.3	**0.0098**	0.004	44.1	ug/L	518	Standard
	Sb	123	500.2	33.4	**0.0994**	0.039	39.6	ug/L	127	Standard
	Ba	135	8786.5	0.8	**4.4621**	0.051	1.1	ug/L	45	Standard
	Ce	140	53.3	19.5				ug/L	298	Standard
>	**Tb**	159	899038.1	0.5				ug/L	977942	Standard
	Ho	165	3.3	86.6				ug/L	12	Standard
	Tl	203	124.3	3.7	**0.0110**	0.001	5.1	ug/L	31	Standard
	Tl	205	108.3	13.3	**0.0136**	0.002	15.1	ug/L	20	Standard
	Pb	206	338.3	5.2	**0.0034**	0.004	109.0	ug/L	286	Standard
	Pb	207	309.3	7.0	**0.0098**	0.005	45.9	ug/L	257	Standard
	Pb	208	1266.0	2.0	**0.0114**	0.001	11.9	ug/L	956	Standard
	U	238	68.0	9.6	**0.0072**	0.001	13.2	ug/L	17	Standard
>	**Bi**	209	520681.9	0.4				ug/L	557080	Standard

Sample ID: L1606073202SDL WG574571-02
Report Date/Time: Tuesday, June 28, 2016 22:07:55
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	5104.2	1.7	**6.8887**	0.170	2.5	mg/L	32	Standard
K	39	40.0	37.5	**-0.0032**	0.052	1630.6	mg/L	30	Standard
Ca	43	30.0	57.7	**-0.6841**	2.704	395.3	mg/L	48	Standard
Fe	54	149.1	25.4	**0.0007**	0.034	4947.2	mg/L	159	Standard
Fe	57	220.0	14.9	**0.1340**	0.108	80.4	mg/L	195	Standard
Sc-1	45	38686.0	0.8				mg/L	39543	Standard
Cl	35	39148.5	1.2				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	4957.5	4.0				ug/L	1647	Standard
P	31	18357.8	2.1				ug/L	15231	Standard
S	34	2318.5	4.4				ug/L	1845	Standard
Sr	88	78.3	18.4				ug/L	70	Standard
C	12	130.0	27.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.2	116.5				mg/L	12	Standard
Ho-1	165	3.3	86.6				mg/L	12	Standard
Er	166	16.7	91.7				mg/L	20	Standard
I	127	32523.3	5.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.021	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.117	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606073202SDL WG574571-02
Report Date/Time: Tuesday, June 28, 2016 22:07:55
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.858
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		93.466
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 22:08:51
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74027.2	2.8				ug/L	81760	Standard
	Be	9	46885.8	2.2	**52.1458**	1.090	2.1	ug/L	17	Standard
	Al	27	7258519.9	1.3	**52.0333**	2.108	4.1	ug/L	1410	Standard
	Sc	45	37711.9	1.9				ug/L	39543	Standard
	Ti	47	24276.8	0.9	**101.7561**	0.915	0.9	ug/L	39	Standard
	V	51	338117.2	0.9	**50.5811**	0.467	0.9	ug/L	1415	Standard
	Cr	52	344759.2	1.1	**50.6515**	0.612	1.2	ug/L	6946	Standard
	Cr	53	43998.8	1.8	**51.9886**	0.950	1.8	ug/L	328	Standard
	Mn	55	319756.8	1.7	**51.3453**	0.844	1.6	ug/L	823	Standard
	Co	59	342187.7	2.1	**50.8146**	1.023	2.0	ug/L	252	Standard
	Ni	60	91463.9	0.8	**50.5986**	0.401	0.8	ug/L	192	Standard
	Cu	65	92539.7	0.2	**50.7558**	0.132	0.3	ug/L	141	Standard
	Zn	66	54475.1	1.3	**51.0616**	0.732	1.4	ug/L	491	Standard
>	**Ge**	72	493368.3	0.1				ug/L	515087	Standard
	As	75	54361.0	0.4	**50.5191**	0.195	0.4	ug/L	20	Standard
	Se	82	5570.4	0.7	**50.3948**	0.327	0.6	ug/L	15	Standard
	Se-1	77	3889.2	2.7	**53.1789**	1.513	2.8	ug/L	48	Standard
>	**Ga**	71	56.7	13.5				mg/L	90	Standard
	Rb	85	710.0	9.2				ug/L	23	Standard
	Y	89	427120.1	2.2				ug/L	448510	Standard
>	**Rh**	103	20.0	50.0				ug/L	15	Standard
	Mo	98	255476.4	1.0	**100.3761**	0.479	0.5	ug/L	11	Standard
	Ag	107	311158.3	0.6	**49.2259**	0.757	1.5	ug/L	134	Standard
	Cd	111	98620.7	1.2	**49.0052**	0.618	1.3	mg/L	7	Standard
	Cd	114	238558.2	0.5	**48.7672**	0.251	0.5	ug/L	16	Standard
>	**In**	115	575337.5	0.9				ug/L	611117	Standard
	Sn	118	277257.2	0.9	**49.7064**	0.586	1.2	ug/L	518	Standard
	Sb	123	223054.4	0.8	**50.2070**	0.812	1.6	ug/L	127	Standard
	Ba	135	94954.3	0.7	**48.8337**	0.782	1.6	ug/L	45	Standard
	Ce	140	61.7	16.9				ug/L	298	Standard
>	**Tb**	159	913449.5	0.5				ug/L	977942	Standard
	Ho	165	10.0	86.6				ug/L	12	Standard
	Tl	203	397249.4	0.6	**50.3639**	0.352	0.7	ug/L	31	Standard
	Tl	205	354959.8	0.3	**51.8245**	0.065	0.1	ug/L	20	Standard
	Pb	206	256290.8	0.6	**50.5971**	0.273	0.5	ug/L	286	Standard
	Pb	207	233188.9	1.2	**51.0418**	0.565	1.1	ug/L	257	Standard
	Pb	208	935979.9	1.1	**52.9561**	0.623	1.2	ug/L	956	Standard
	U	238	344850.7	0.5	**51.4151**	0.322	0.6	ug/L	17	Standard
>	**Bi**	209	521458.9	0.2				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:11:08
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3802.1	3.0	**5.2525**	0.237	4.5	mg/L	32	Standard
K	39	1311.7	9.2	**4.5102**	0.493	10.9	mg/L	30	Standard
Ca	43	78.3	7.4	**7.1225**	0.723	10.1	mg/L	48	Standard
Fe	54	5407.8	4.5	**4.9980**	0.312	6.2	mg/L	159	Standard
Fe	57	1643.4	4.4	**5.1911**	0.156	3.0	mg/L	195	Standard
> **Sc-1**	45	37711.9	1.9				mg/L	39543	Standard
Cl	35	36184.5	2.9				ug/L	27642	Standard
Kr	83	3.0	66.7				ug/L	2	Standard
Br	81	1690.1	2.4				ug/L	1647	Standard
P	31	19537.6	2.2				ug/L	15231	Standard
S	34	2665.2	5.3				ug/L	1845	Standard
Sr	88	80.0	12.5				ug/L	70	Standard
C	12	116.7	9.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	9.4	109.4				mg/L	12	Standard
Ho-1	165	10.0	86.6				mg/L	12	Standard
Er	166	13.3	43.3				mg/L	20	Standard
I	127	5821.1	15.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	104.292		
	Al	27	104.067		
	Sc	45			
	Ti	47	101.756		
	V	51	101.162		
	Cr	52	101.303		
	Cr	53			
	Mn	55	102.691		
	Co	59	101.629		
	Ni	60	101.197		
	Cu	65	101.512		
	Zn	66	102.123		
>	Ge	72		95.784	
	As	75	101.038		
	Se	82	100.790		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:11:08
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	100.376		
Ag	107	98.452		
Cd	111	98.010		
Cd	114			
> In	115			94.145
Sn	118	99.413		
Sb	123	100.414		
Ba	135	97.667		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	100.728		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.912		
U	238	102.830		
> Bi	209			93.606
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:11:08
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 22:12:02
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72720.8	1.4				ug/L	81760	Standard
	Be	9	31.7	9.1	**0.0261**	0.003	13.4	ug/L	17	Standard
	Al	27	3567.1	29.0	**0.0117**	0.008	67.2	ug/L	1410	Standard
	Sc	45	39132.1	1.1				ug/L	39543	Standard
	Ti	47	37.0	21.5	**-0.0116**	0.034	290.1	ug/L	39	Standard
	V	51	1125.9	7.5	**-0.0395**	0.013	33.1	ug/L	1415	Standard
	Cr	52	6280.0	4.7	**-0.1269**	0.043	33.9	ug/L	6946	Standard
	Cr	53	665.0	3.8	**0.4397**	0.027	6.1	ug/L	328	Standard
	Mn	55	716.0	7.9	**-0.0346**	0.009	26.9	ug/L	823	Standard
	Co	59	261.7	5.4	**0.0070**	0.002	26.9	ug/L	252	Standard
	Ni	60	220.7	25.5	**0.0158**	0.031	194.2	ug/L	192	Standard
	Cu	65	175.7	31.3	**0.0045**	0.030	664.0	ug/L	141	Standard
	Zn	66	494.7	12.2	**-0.3623**	0.056	15.6	ug/L	491	Standard
>	**Ge**	72	499522.9	0.5				ug/L	515087	Standard
	As	75	-22.4	199.4	**0.0197**	0.041	208.7	ug/L	20	Standard
	Se	82	18.1	44.1	**0.0430**	0.071	165.8	ug/L	15	Standard
	Se-1	77	56.7	13.4	**0.0566**	0.102	180.9	ug/L	48	Standard
>	**Ga**	71	45.0	38.5				mg/L	90	Standard
	Rb	85	31.7	45.6				ug/L	23	Standard
	Y	89	417360.5	0.9				ug/L	448510	Standard
>	**Rh**	103	6.7	114.6				ug/L	15	Standard
	Mo	98	287.6	23.9	**0.0927**	0.026	27.9	ug/L	11	Standard
	Ag	107	190.3	4.8	**0.0136**	0.002	11.9	ug/L	134	Standard
	Cd	111	25.8	62.0	**0.0098**	0.008	81.4	mg/L	7	Standard
	Cd	114	75.5	64.7	**0.0104**	0.010	96.5	ug/L	16	Standard
>	**In**	115	581386.0	1.0				ug/L	611117	Standard
	Sn	118	788.4	29.5	**0.0382**	0.041	108.2	ug/L	518	Standard
	Sb	123	553.2	30.8	**0.1105**	0.038	34.6	ug/L	127	Standard
	Ba	135	104.3	62.1	**0.0282**	0.034	119.2	ug/L	45	Standard
	Ce	140	15.0	88.2				ug/L	298	Standard
>	**Tb**	159	916886.1	2.0				ug/L	977942	Standard
	Ho	165	6.7	43.3				ug/L	12	Standard
	Tl	203	96.0	40.0	**0.0072**	0.005	66.3	ug/L	31	Standard
	Tl	205	80.0	25.0	**0.0092**	0.003	29.7	ug/L	20	Standard
	Pb	206	283.3	0.7	**-0.0085**	0.001	14.2	ug/L	286	Standard
	Pb	207	242.3	7.5	**-0.0058**	0.004	64.9	ug/L	257	Standard
	Pb	208	1066.3	4.7	**-0.0011**	0.002	176.0	ug/L	956	Standard
	U	238	68.3	24.9	**0.0070**	0.002	33.1	ug/L	17	Standard
>	**Bi**	209	530213.7	1.5				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:14:19
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**32.2313**	32.195	99.9	mg/L	0	Standard
Mg	24	43.3	46.6	**0.0012**	0.027	2331.0	mg/L	32	Standard
K	39	38.3	41.9	**-0.0103**	0.056	538.4	mg/L	30	Standard
Ca	43	25.0	20.0	**-1.5137**	0.732	48.3	mg/L	48	Standard
Fe	54	136.0	18.8	**-0.0125**	0.025	197.8	mg/L	159	Standard
Fe	57	235.0	17.0	**0.1764**	0.130	73.9	mg/L	195	Standard
Sc-1	45	39132.1	1.1				mg/L	39543	Standard
Cl	35	35003.7	1.3				ug/L	27642	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	1530.1	14.2				ug/L	1647	Standard
P	31	18486.3	1.6				ug/L	15231	Standard
S	34	2496.9	3.2				ug/L	1845	Standard
Sr	88	51.7	31.1				ug/L	70	Standard
C	12	86.7	52.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.5	5.0				mg/L	12	Standard
Ho-1	165	6.7	43.3				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	4170.6	14.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.978	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:14:19
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.135
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.177
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:14:19
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606073202
Sample Date/Time: Tuesday, June 28, 2016 22:15:14
Number of Replicates: 3
Autosampler Position: 338
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76318.4	2.0				ug/L	81760	Standard
	Be	9	40.0	69.6	**0.0330**	0.029	88.9	ug/L	17	Standard
	Al	27	23571551.0	1.4	**163.8296**	2.147	1.3	ug/L	1410	Standard
	Sc	45	39020.2	3.0				ug/L	39543	Standard
	Ti	47	595.0	5.6	**2.3771**	0.145	6.1	ug/L	39	Standard
	V	51	1578.4	25.4	**0.0346**	0.061	176.4	ug/L	1415	Standard
	Cr	52	21713.5	1.1	**2.2571**	0.032	1.4	ug/L	6946	Standard
	Cr	53	25227.7	4.6	**30.1930**	1.400	4.6	ug/L	328	Standard
	Mn	55	113835.4	1.3	**18.5078**	0.247	1.3	ug/L	823	Standard
	Co	59	3991.2	2.2	**0.5722**	0.015	2.6	ug/L	252	Standard
	Ni	60	3905.8	1.2	**2.0985**	0.020	1.0	ug/L	192	Standard
	Cu	65	1542.7	2.9	**0.7718**	0.023	3.0	ug/L	141	Standard
	Zn	66	25635.3	1.1	**24.0224**	0.322	1.3	ug/L	491	Standard
>	**Ge**	72	484802.3	0.3				ug/L	515087	Standard
	As	75	911.4	10.9	**0.9016**	0.094	10.4	ug/L	20	Standard
	Se	82	285.6	1.4	**2.5161**	0.032	1.3	ug/L	15	Standard
	Se-1	77	1182.4	4.2	**15.9563**	0.701	4.4	ug/L	48	Standard
>	**Ga**	71	88.3	32.7				mg/L	90	Standard
	Rb	85	8777.5	4.2				ug/L	23	Standard
	Y	89	401864.9	0.9				ug/L	448510	Standard
>	**Rh**	103	95.0	19.0				ug/L	15	Standard
	Mo	98	789.3	4.8	**0.3176**	0.018	5.8	ug/L	11	Standard
	Ag	107	101.3	14.3	**0.0012**	0.002	197.6	ug/L	134	Standard
	Cd	111	81.6	8.8	**0.0411**	0.003	8.4	mg/L	7	Standard
	Cd	114	236.0	13.0	**0.0475**	0.007	15.0	ug/L	16	Standard
>	**In**	115	530144.7	1.0				ug/L	611117	Standard
	Sn	118	770.0	10.1	**0.0481**	0.014	28.6	ug/L	518	Standard
	Sb	123	728.9	13.4	**0.1651**	0.022	13.6	ug/L	127	Standard
	Ba	135	185424.0	0.3	**103.5140**	0.966	0.9	ug/L	45	Standard
	Ce	140	496.7	18.3				ug/L	298	Standard
>	**Tb**	159	867434.9	1.9				ug/L	977942	Standard
	Ho	165	20.0	0.0				ug/L	12	Standard
	Tl	203	239.3	4.0	**0.0278**	0.001	4.3	ug/L	31	Standard
	Tl	205	225.0	20.4	**0.0330**	0.007	22.5	ug/L	20	Standard
	Pb	206	582.0	1.5	**0.0598**	0.001	1.4	ug/L	286	Standard
	Pb	207	480.3	3.9	**0.0549**	0.005	9.4	ug/L	257	Standard
	Pb	208	1935.7	1.6	**0.0572**	0.003	4.8	ug/L	956	Standard
	U	238	1057.0	6.3	**0.1661**	0.010	5.7	ug/L	17	Standard
>	**Bi**	209	486031.1	0.8				ug/L	557080	Standard

Sample ID: L1606073202
Report Date/Time: Tuesday, June 28, 2016 22:17:31
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	123145.1	1.2	**166.1283**	3.309	2.0	mg/L	32	Standard
K	39	193.3	27.9	**0.5225**	0.193	37.0	mg/L	30	Standard
Ca	43	155.0	19.6	**18.5842**	5.220	28.1	mg/L	48	Standard
Fe	54	201.4	15.9	**0.0482**	0.034	70.8	mg/L	159	Standard
Fe	57	301.7	7.8	**0.4095**	0.108	26.4	mg/L	195	Standard
> **Sc-1**	45	39020.2	3.0				mg/L	39543	Standard
Cl	35	36882.8	1.2				ug/L	27642	Standard
Kr	83	2.3	65.5				ug/L	2	Standard
Br	81	66321.6	2.9				ug/L	1647	Standard
P	31	23950.6	0.8				ug/L	15231	Standard
S	34	2285.2	3.2				ug/L	1845	Standard
Sr	88	326.7	6.4				ug/L	70	Standard
C	12	383.3	11.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	24.1	87.6				mg/L	12	Standard
Ho-1	165	20.0	0.0				mg/L	12	Standard
Er	166	53.3	28.6				mg/L	20	Standard
I	127	490270.4	8.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.344	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.121	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606073202
Report Date/Time: Tuesday, June 28, 2016 22:17:31
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	86.750
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	87.246
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606073202
Report Date/Time: Tuesday, June 28, 2016 22:17:31
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606076201
Sample Date/Time: Tuesday, June 28, 2016 22:18:25
Number of Replicates: 3
Autosampler Position: 339
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74725.5	1.6				ug/L	81760	Standard
	Be	9	33.3	56.8	**0.0272**	0.022	79.2	ug/L	17	Standard
	Al	27	1272215.4	1.6	**9.0180**	0.230	2.5	ug/L	1410	Standard
	Sc	45	40079.6	0.9				ug/L	39543	Standard
	Ti	47	73.0	21.4	**0.1312**	0.065	49.6	ug/L	39	Standard
	V	51	1912.1	11.7	**0.0708**	0.033	46.3	ug/L	1415	Standard
	Cr	52	9243.8	1.7	**0.2820**	0.015	5.4	ug/L	6946	Standard
	Cr	53	1550.1	16.6	**1.4381**	0.284	19.7	ug/L	328	Standard
	Mn	55	4050.6	51.7	**0.4818**	0.327	67.8	ug/L	823	Standard
	Co	59	576.0	42.1	**0.0513**	0.035	68.0	ug/L	252	Standard
	Ni	60	659.3	6.6	**0.2481**	0.023	9.5	ug/L	192	Standard
	Cu	65	493.7	5.0	**0.1712**	0.013	7.5	ug/L	141	Standard
	Zn	66	2294.8	5.3	**1.2839**	0.114	8.9	ug/L	491	Standard
>	**Ge**	72	510757.4	0.7				ug/L	515087	Standard
	As	75	-15.4	127.0	**0.0265**	0.018	66.6	ug/L	20	Standard
	Se	82	23.9	38.0	**0.0895**	0.079	87.8	ug/L	15	Standard
	Se-1	77	93.7	16.1	**0.5342**	0.193	36.2	ug/L	48	Standard
>	**Ga**	71	58.3	17.8				mg/L	90	Standard
	Rb	85	1476.7	6.6				ug/L	23	Standard
	Y	89	432514.2	1.7				ug/L	448510	Standard
>	**Rh**	103	26.7	39.0				ug/L	15	Standard
	Mo	98	146.0	48.2	**0.0371**	0.027	72.5	ug/L	11	Standard
	Ag	107	112.7	13.4	**0.0013**	0.002	181.1	ug/L	134	Standard
	Cd	111	43.2	14.9	**0.0182**	0.003	18.1	mg/L	7	Standard
	Cd	114	109.3	34.7	**0.0170**	0.007	43.8	ug/L	16	Standard
>	**In**	115	586396.8	0.9				ug/L	611117	Standard
	Sn	118	771.7	15.6	**0.0341**	0.022	63.3	ug/L	518	Standard
	Sb	123	236.4	15.7	**0.0394**	0.008	21.0	ug/L	127	Standard
	Ba	135	3173.0	2.9	**1.5769**	0.056	3.5	ug/L	45	Standard
	Ce	140	2270.2	3.3				ug/L	298	Standard
>	**Tb**	159	916697.6	0.7				ug/L	977942	Standard
	Ho	165	490.0	12.4				ug/L	12	Standard
	Tl	203	130.7	10.3	**0.0114**	0.002	13.6	ug/L	31	Standard
	Tl	205	215.0	77.4	**0.0282**	0.023	82.9	ug/L	20	Standard
	Pb	206	395.3	3.6	**0.0126**	0.003	22.5	ug/L	286	Standard
	Pb	207	368.7	19.2	**0.0207**	0.015	70.3	ug/L	257	Standard
	Pb	208	1630.7	27.1	**0.0295**	0.024	80.4	ug/L	956	Standard
	U	238	158.7	69.5	**0.0200**	0.016	79.0	ug/L	17	Standard
>	**Bi**	209	534609.1	0.7				ug/L	557080	Standard

Sample ID: L1606076201
Report Date/Time: Tuesday, June 28, 2016 22:20:42
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**31.5470**	31.355	99.4	mg/L	0	Standard
Mg	24	598.3	11.5	**0.7287**	0.091	12.5	mg/L	32	Standard
K	39	176.7	18.2	**0.4475**	0.107	24.0	mg/L	30	Standard
Ca	43	93.3	48.6	**8.6233**	6.782	78.6	mg/L	48	Standard
Fe	54	114.2	15.8	**-0.0351**	0.016	46.3	mg/L	159	Standard
Fe	57	271.7	7.4	**0.2806**	0.074	26.3	mg/L	195	Standard
> **Sc-1**	45	40079.6	0.9				mg/L	39543	Standard
Cl	35	37059.3	1.9				ug/L	27642	Standard
Kr	83	2.0	86.6				ug/L	2	Standard
Br	81	3193.7	9.3				ug/L	1647	Standard
P	31	19133.7	1.0				ug/L	15231	Standard
S	34	2435.2	3.0				ug/L	1845	Standard
Sr	88	116.7	19.8				ug/L	70	Standard
C	12	110.0	24.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	699.7	7.7				mg/L	12	Standard
Ho-1	165	490.0	12.4				mg/L	12	Standard
Er	166	563.3	4.5				mg/L	20	Standard
I	127	24296.3	37.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.396	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.159	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606076201
Report Date/Time: Tuesday, June 28, 2016 22:20:42
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.955
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.966
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606138001
Sample Date/Time: Tuesday, June 28, 2016 22:21:37
Number of Replicates: 3
Autosampler Position: 340
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79317.1	2.1				ug/L	81760	Standard
	Be	9	923.4	3.6	**0.9491**	0.046	4.8	ug/L	17	Standard
	Al	27	22322355.7	1.2	**149.2842**	1.816	1.2	ug/L	1410	Standard
	Sc	45	50147.9	2.6				ug/L	39543	Standard
	Ti	47	26223.7	0.7	**107.9697**	0.846	0.8	ug/L	39	Standard
	V	51	127306.3	1.5	**18.5751**	0.291	1.6	ug/L	1415	Standard
	Cr	52	100654.9	1.1	**13.7705**	0.145	1.1	ug/L	6946	Standard
	Cr	53	12865.5	4.5	**14.6884**	0.707	4.8	ug/L	328	Standard
	Mn	55	4304305.9	0.4	**680.6871**	4.381	0.6	ug/L	823	Standard
	Co	59	49359.3	0.1	**7.1725**	0.024	0.3	ug/L	252	Standard
	Ni	60	30538.1	1.5	**16.5228**	0.280	1.7	ug/L	192	Standard
	Cu	65	24616.6	0.6	**13.1942**	0.101	0.8	ug/L	141	Standard
	Zn	66	126791.0	1.1	**117.7932**	1.321	1.1	ug/L	491	Standard
>	**Ge**	72	502314.8	0.3				ug/L	515087	Standard
	As	75	6358.7	2.6	**5.8397**	0.152	2.6	ug/L	20	Standard
	Se	82	52.0	15.4	**0.3436**	0.072	20.9	ug/L	15	Standard
	Se-1	77	111.0	6.5	**0.7919**	0.097	12.2	ug/L	48	Standard
>	**Ga**	71	15155.9	0.2				mg/L	90	Standard
	Rb	85	244761.5	1.6				ug/L	23	Standard
	Y	89	581808.2	2.1				ug/L	448510	Standard
>	**Rh**	103	81.7	28.3				ug/L	15	Standard
	Mo	98	7400.1	1.4	**2.8571**	0.099	3.5	ug/L	11	Standard
	Ag	107	359.7	8.1	**0.0401**	0.005	12.7	ug/L	134	Standard
	Cd	111	226.0	4.7	**0.1082**	0.006	6.0	mg/L	7	Standard
	Cd	114	572.5	10.7	**0.1110**	0.014	12.9	ug/L	16	Standard
>	**In**	115	581977.2	2.1				ug/L	611117	Standard
	Sn	118	1150.0	8.4	**0.1021**	0.013	12.6	ug/L	518	Standard
	Sb	123	1041.4	3.0	**0.2191**	0.011	5.1	ug/L	127	Standard
	Ba	135	264934.9	0.4	**134.7601**	2.267	1.7	ug/L	45	Standard
	Ce	140	872470.7	0.6				ug/L	298	Standard
>	**Tb**	159	944063.2	1.7				ug/L	977942	Standard
	Ho	165	13182.4	3.3				ug/L	12	Standard
	Tl	203	1500.7	3.5	**0.1808**	0.008	4.2	ug/L	31	Standard
	Tl	205	1268.4	3.4	**0.1783**	0.007	4.2	ug/L	20	Standard
	Pb	206	85066.5	1.3	**16.3322**	0.328	2.0	ug/L	286	Standard
	Pb	207	70090.8	1.2	**14.9184**	0.287	1.9	ug/L	257	Standard
	Pb	208	293052.1	0.7	**16.1237**	0.113	0.7	ug/L	956	Standard
	U	238	6791.5	1.5	**0.9844**	0.021	2.1	ug/L	17	Standard
>	**Bi**	209	534845.3	0.8				ug/L	557080	Standard

Sample ID: L1606138001
Report Date/Time: Tuesday, June 28, 2016 22:23:54
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	2491.9	2.3	**2.5590**	0.057	2.2	mg/L	32	Standard
K	39	860.0	5.6	**2.1528**	0.183	8.5	mg/L	30	Standard
Ca	43	481.7	3.2	**52.3892**	3.048	5.8	mg/L	48	Standard
Fe	54	14060.9	4.3	**9.8944**	0.202	2.0	mg/L	159	Standard
Fe	57	4057.2	4.0	**10.1835**	0.599	5.9	mg/L	195	Standard
> **Sc-1**	45	50147.9	2.6				mg/L	39543	Standard
Cl	35	34061.6	1.3				ug/L	27642	Standard
Kr	83	2.7	57.3				ug/L	2	Standard
Br	81	3443.7	4.1				ug/L	1647	Standard
P	31	33871.2	1.5				ug/L	15231	Standard
S	34	2461.9	2.1				ug/L	1845	Standard
Sr	88	305.0	8.2				ug/L	70	Standard
C	12	206.7	29.6				mg/L	113	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	40.0	50.0				mg/L	0	Standard
Dy	164	22121.1	1.8				mg/L	12	Standard
Ho-1	165	13182.4	3.3				mg/L	12	Standard
Er	166	12021.4	0.9				mg/L	20	Standard
I	127	158658.4	11.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		97.012	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.520	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138001
Report Date/Time: Tuesday, June 28, 2016 22:23:54
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	95.232
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	96.009
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Ti 47 Upper, S, EEE	Ti	47	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606138001
Report Date/Time: Tuesday, June 28, 2016 22:23:54
Page 3

Approved: June 29, 2016

Zn 66 Upper, S, EEE	Zn	66
Ba 135 Upper, S, EEE	Ba	135

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606138002
Sample Date/Time: Tuesday, June 28, 2016 22:24:48
Number of Replicates: 3
Autosampler Position: 341
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80508.4	1.6				ug/L	81760	Standard
	Be	9	865.0	8.5	**0.8749**	0.075	8.6	ug/L	17	Standard
	Al	27	22868958.8	1.7	**150.6704**	2.885	1.9	ug/L	1410	Standard
	Sc	45	49400.4	2.8				ug/L	39543	Standard
	Ti	47	23164.7	1.4	**96.5938**	1.528	1.6	ug/L	39	Standard
	V	51	124883.8	1.8	**18.4559**	0.304	1.6	ug/L	1415	Standard
	Cr	52	103466.8	1.3	**14.3825**	0.239	1.7	ug/L	6946	Standard
	Cr	53	12758.7	1.5	**14.7559**	0.244	1.7	ug/L	328	Standard
	Mn	55	4407919.2	0.3	**706.0993**	3.755	0.5	ug/L	823	Standard
	Co	59	51228.1	1.3	**7.5421**	0.111	1.5	ug/L	252	Standard
	Ni	60	30969.0	1.2	**16.9756**	0.235	1.4	ug/L	192	Standard
	Cu	65	25034.7	0.7	**13.5947**	0.118	0.9	ug/L	141	Standard
	Zn	66	152662.0	1.2	**143.8470**	1.906	1.3	ug/L	491	Standard
>	**Ge**	72	495895.8	0.2				ug/L	515087	Standard
	As	75	6401.6	1.0	**5.9544**	0.051	0.9	ug/L	20	Standard
	Se	82	65.6	3.6	**0.4728**	0.020	4.3	ug/L	15	Standard
	Se-1	77	120.3	12.9	**0.9403**	0.215	22.9	ug/L	48	Standard
>	**Ga**	71	15282.7	1.5				mg/L	90	Standard
	Rb	85	248664.6	2.3				ug/L	23	Standard
	Y	89	598252.6	1.6				ug/L	448510	Standard
>	**Rh**	103	61.7	4.7				ug/L	15	Standard
	Mo	98	8173.4	1.7	**3.0442**	0.032	1.1	ug/L	11	Standard
	Ag	107	334.0	5.8	**0.0342**	0.003	8.9	ug/L	134	Standard
	Cd	111	175.2	4.5	**0.0802**	0.004	5.0	mg/L	7	Standard
	Cd	114	424.7	8.5	**0.0779**	0.007	8.4	ug/L	16	Standard
>	**In**	115	603234.5	0.7				ug/L	611117	Standard
	Sn	118	1006.7	2.2	**0.0706**	0.004	6.0	ug/L	518	Standard
	Sb	123	981.2	4.8	**0.1979**	0.011	5.8	ug/L	127	Standard
	Ba	135	265403.3	0.1	**130.2141**	0.915	0.7	ug/L	45	Standard
	Ce	140	904460.6	1.2				ug/L	298	Standard
>	**Tb**	159	953156.2	1.3				ug/L	977942	Standard
	Ho	165	13287.5	2.7				ug/L	12	Standard
	Tl	203	1338.1	1.2	**0.1623**	0.004	2.2	ug/L	31	Standard
	Tl	205	1380.1	21.0	**0.1966**	0.045	23.0	ug/L	20	Standard
	Pb	206	85143.9	1.2	**16.5139**	0.424	2.6	ug/L	286	Standard
	Pb	207	70337.2	0.8	**15.1235**	0.335	2.2	ug/L	257	Standard
	Pb	208	299122.5	0.5	**16.6275**	0.345	2.1	ug/L	956	Standard
	U	238	7037.6	3.1	**1.0308**	0.048	4.7	ug/L	17	Standard
>	**Bi**	209	529543.6	1.6				ug/L	557080	Standard

Sample ID: L1606138002
Report Date/Time: Tuesday, June 28, 2016 22:27:05
Page 1

Approved: June 29, 2016

Na	23	5.0	0.0	**25.7128**	0.720	2.8	mg/L	0	Standard
Mg	24	2751.9	1.6	**2.8760**	0.054	1.9	mg/L	32	Standard
K	39	820.0	12.4	**2.0772**	0.274	13.2	mg/L	30	Standard
Ca	43	460.0	15.1	**50.7618**	10.089	19.9	mg/L	48	Standard
Fe	54	14800.2	1.5	**10.5924**	0.445	4.2	mg/L	159	Standard
Fe	57	4213.9	2.7	**10.7740**	0.623	5.8	mg/L	195	Standard
Sc-1	45	49400.4	2.8				mg/L	39543	Standard
Cl	35	34900.8	0.4				ug/L	27642	Standard
Kr	83	3.0	0.0				ug/L	2	Standard
Br	81	3303.7	3.6				ug/L	1647	Standard
P	31	33101.1	2.4				ug/L	15231	Standard
S	34	2478.5	3.3				ug/L	1845	Standard
Sr	88	288.3	13.9				ug/L	70	Standard
C	12	206.7	17.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	53.3	65.8				mg/L	0	Standard
Dy	164	23538.0	3.9				mg/L	12	Standard
Ho-1	165	13287.5	2.7				mg/L	12	Standard
Er	166	12271.6	4.5				mg/L	20	Standard
I	127	174817.7	8.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		98.469	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.274	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138002
Report Date/Time: Tuesday, June 28, 2016 22:27:05
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.710
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.057
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606138002
Report Date/Time: Tuesday, June 28, 2016 22:27:05
Page 3

Approved: June 29, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606138106
Sample Date/Time: Tuesday, June 28, 2016 22:28:00
Number of Replicates: 3
Autosampler Position: 342
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75884.6	2.0				ug/L	81760	Standard
	Be	9	26.7	10.8	**0.0192**	0.003	17.5	ug/L	17	Standard
	Al	27	26057.5	7.4	**0.1676**	0.010	6.0	ug/L	1410	Standard
	Sc	45	39473.0	0.8				ug/L	39543	Standard
	Ti	47	39.3	21.6	**-0.0051**	0.035	681.9	ug/L	39	Standard
	V	51	1622.1	7.1	**0.0295**	0.017	56.8	ug/L	1415	Standard
	Cr	52	8616.4	0.2	**0.1952**	0.003	1.5	ug/L	6946	Standard
	Cr	53	775.0	5.7	**0.5518**	0.051	9.2	ug/L	328	Standard
	Mn	55	5073.2	5.2	**0.6432**	0.042	6.5	ug/L	823	Standard
	Co	59	305.3	2.9	**0.0126**	0.001	10.1	ug/L	252	Standard
	Ni	60	336.7	8.9	**0.0759**	0.016	21.2	ug/L	192	Standard
	Cu	65	1953.8	0.6	**0.9495**	0.006	0.6	ug/L	141	Standard
	Zn	66	2327.2	0.6	**1.3204**	0.013	1.0	ug/L	491	Standard
>	**Ge**	72	509122.8	0.0				ug/L	515087	Standard
	As	75	-29.7	15.9	**0.0136**	0.004	31.1	ug/L	20	Standard
	Se	82	19.1	15.2	**0.0489**	0.026	52.3	ug/L	15	Standard
	Se-1	77	54.3	12.9	**0.0108**	0.094	874.4	ug/L	48	Standard
>	**Ga**	71	61.7	23.4				mg/L	90	Standard
	Rb	85	193.3	9.8				ug/L	23	Standard
	Y	89	438080.9	2.7				ug/L	448510	Standard
>	**Rh**	103	13.3	94.4				ug/L	15	Standard
	Mo	98	1348.1	2.9	**0.4917**	0.018	3.6	ug/L	11	Standard
	Ag	107	96.0	6.3	**-0.0016**	0.001	59.2	ug/L	134	Standard
	Cd	111	29.0	10.6	**0.0110**	0.001	13.2	mg/L	7	Standard
	Cd	114	62.4	58.3	**0.0074**	0.007	97.2	ug/L	16	Standard
>	**In**	115	596825.6	0.5				ug/L	611117	Standard
	Sn	118	1153.4	11.3	**0.0978**	0.023	23.4	ug/L	518	Standard
	Sb	123	161.5	18.7	**0.0222**	0.007	29.7	ug/L	127	Standard
	Ba	135	505.0	6.7	**0.2253**	0.016	7.0	ug/L	45	Standard
	Ce	140	233.3	21.5				ug/L	298	Standard
>	**Tb**	159	918727.2	0.8				ug/L	977942	Standard
	Ho	165	6.7	43.3				ug/L	12	Standard
	Tl	203	77.0	15.3	**0.0048**	0.001	30.3	ug/L	31	Standard
	Tl	205	70.0	7.1	**0.0078**	0.001	10.7	ug/L	20	Standard
	Pb	206	353.0	8.2	**0.0052**	0.006	124.2	ug/L	286	Standard
	Pb	207	316.0	5.0	**0.0102**	0.004	41.5	ug/L	257	Standard
	Pb	208	1279.0	2.9	**0.0109**	0.002	20.1	ug/L	956	Standard
	U	238	66.0	9.1	**0.0067**	0.001	13.4	ug/L	17	Standard
>	**Bi**	209	529518.6	1.3				ug/L	557080	Standard

Sample ID: L1606138106
Report Date/Time: Tuesday, June 28, 2016 22:30:17
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.6931**	18.512	173.1	mg/L	0	Standard
Mg	24	105.0	31.2	**0.0830**	0.044	52.9	mg/L	32	Standard
K	39	31.7	9.1	**-0.0342**	0.011	31.1	mg/L	30	Standard
Ca	43	30.0	66.7	**-0.7662**	3.081	402.2	mg/L	48	Standard
Fe	54	132.0	13.9	**-0.0174**	0.016	93.3	mg/L	159	Standard
Fe	57	243.3	16.5	**0.1990**	0.142	71.4	mg/L	195	Standard
Sc-1	45	39473.0	0.8				mg/L	39543	Standard
Cl	35	35269.0	1.4				ug/L	27642	Standard
Kr	83	2.3	107.9				ug/L	2	Standard
Br	81	2306.8	8.4				ug/L	1647	Standard
P	31	16594.1	0.5				ug/L	15231	Standard
S	34	2131.8	6.4				ug/L	1845	Standard
Sr	88	88.3	36.8				ug/L	70	Standard
C	12	163.3	23.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	0	Standard
Dy	164	9.0	110.6				mg/L	12	Standard
Ho-1	165	6.7	43.3				mg/L	12	Standard
Er	166	20.0	0.0				mg/L	20	Standard
I	127	8462.7	47.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		92.814	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		98.842	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138106
Report Date/Time: Tuesday, June 28, 2016 22:30:17
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.661
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.052
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139301
Sample Date/Time: Tuesday, June 28, 2016 22:31:11
Number of Replicates: 3
Autosampler Position: 343
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75653.4	0.9				ug/L	81760	Standard
	Be	9	36.7	15.7	**0.0301**	0.006	19.8	ug/L	17	Standard
	Al	27	730309.3	1.1	**5.1065**	0.100	1.9	ug/L	1410	Standard
	Sc	45	40455.6	1.8				ug/L	39543	Standard
	Ti	47	3049.6	0.8	**12.0214**	0.208	1.7	ug/L	39	Standard
	V	51	20044.5	0.4	**2.6590**	0.030	1.1	ug/L	1415	Standard
	Cr	52	14783.9	1.2	**1.0533**	0.031	2.9	ug/L	6946	Standard
	Cr	53	1555.1	3.4	**1.4185**	0.043	3.0	ug/L	328	Standard
	Mn	55	57374.1	1.8	**8.6446**	0.070	0.8	ug/L	823	Standard
	Co	59	1706.1	0.5	**0.2099**	0.002	1.1	ug/L	252	Standard
	Ni	60	1958.1	2.2	**0.9282**	0.028	3.1	ug/L	192	Standard
	Cu	65	11994.7	1.9	**6.1812**	0.063	1.0	ug/L	141	Standard
	Zn	66	20161.1	0.6	**17.4500**	0.168	1.0	ug/L	491	Standard
>	**Ge**	72	518413.1	1.4				ug/L	515087	Standard
	As	75	157.8	25.1	**0.1801**	0.037	20.6	ug/L	20	Standard
	Se	82	23.1	48.8	**0.0801**	0.098	121.9	ug/L	15	Standard
	Se-1	77	63.3	7.3	**0.1169**	0.069	59.2	ug/L	48	Standard
>	**Ga**	71	505.0	4.5				mg/L	90	Standard
	Rb	85	7842.0	2.0				ug/L	23	Standard
	Y	89	431479.5	0.9				ug/L	448510	Standard
>	**Rh**	103	15.0	100.0				ug/L	15	Standard
	Mo	98	468.3	2.4	**0.1592**	0.005	3.4	ug/L	11	Standard
	Ag	107	177.3	7.3	**0.0110**	0.002	18.6	ug/L	134	Standard
	Cd	111	203.8	5.3	**0.0952**	0.005	5.7	mg/L	7	Standard
	Cd	114	524.4	13.4	**0.0989**	0.013	13.5	ug/L	16	Standard
>	**In**	115	593880.8	0.6				ug/L	611117	Standard
	Sn	118	2320.2	4.6	**0.3018**	0.016	5.3	ug/L	518	Standard
	Sb	123	1520.6	3.7	**0.3188**	0.011	3.4	ug/L	127	Standard
	Ba	135	21231.5	1.4	**10.5571**	0.086	0.8	ug/L	45	Standard
	Ce	140	21453.5	1.1				ug/L	298	Standard
>	**Tb**	159	941666.6	1.6				ug/L	977942	Standard
	Ho	165	511.7	11.3				ug/L	12	Standard
	Tl	203	106.0	9.9	**0.0082**	0.001	16.0	ug/L	31	Standard
	Tl	205	85.0	31.1	**0.0097**	0.004	38.6	ug/L	20	Standard
	Pb	206	17358.9	1.4	**3.2466**	0.044	1.3	ug/L	286	Standard
	Pb	207	14395.2	1.0	**2.9852**	0.035	1.2	ug/L	257	Standard
	Pb	208	59327.6	0.5	**3.1814**	0.019	0.6	ug/L	956	Standard
	U	238	205.3	8.1	**0.0266**	0.002	9.0	ug/L	17	Standard
>	**Bi**	209	540531.5	0.2				ug/L	557080	Standard

Sample ID: L1606139301
Report Date/Time: Tuesday, June 28, 2016 22:33:28
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**21.0062**	36.375	173.2	mg/L	0	Standard
Mg	24	250.0	13.1	**0.2677**	0.037	13.9	mg/L	32	Standard
K	39	110.0	28.4	**0.2220**	0.104	47.1	mg/L	30	Standard
Ca	43	51.7	20.1	**2.3448**	1.674	71.4	mg/L	48	Standard
Fe	54	595.3	12.6	**0.3888**	0.057	14.6	mg/L	159	Standard
Fe	57	388.3	11.5	**0.6585**	0.168	25.5	mg/L	195	Standard
> **Sc-1**	45	40455.6	1.8				mg/L	39543	Standard
Cl	35	34669.0	0.4				ug/L	27642	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	3693.8	7.2				ug/L	1647	Standard
P	31	17358.3	1.4				ug/L	15231	Standard
S	34	2180.2	7.0				ug/L	1845	Standard
Sr	88	80.0	38.0				ug/L	70	Standard
C	12	186.7	21.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	0	Standard
Dy	164	723.1	8.6				mg/L	12	Standard
Ho-1	165	511.7	11.3				mg/L	12	Standard
Er	166	493.3	12.2				mg/L	20	Standard
I	127	46309.3	8.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		92.531	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.646	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606139301
Report Date/Time: Tuesday, June 28, 2016 22:33:28
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.180
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.029
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139302
Sample Date/Time: Tuesday, June 28, 2016 22:34:23
Number of Replicates: 3
Autosampler Position: 344
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76487.6	0.6				ug/L	81760	Standard
	Be	9	26.7	57.3	**0.0190**	0.016	86.6	ug/L	17	Standard
	Al	27	422387.9	1.8	**2.9148**	0.060	2.0	ug/L	1410	Standard
	Sc	45	40846.7	1.8				ug/L	39543	Standard
	Ti	47	1809.8	2.0	**7.1474**	0.208	2.9	ug/L	39	Standard
	V	51	13250.7	0.8	**1.7088**	0.032	1.8	ug/L	1415	Standard
	Cr	52	14384.8	1.8	**1.0188**	0.008	0.8	ug/L	6946	Standard
	Cr	53	1510.1	5.3	**1.3859**	0.061	4.4	ug/L	328	Standard
	Mn	55	44577.2	1.1	**6.7600**	0.108	1.6	ug/L	823	Standard
	Co	59	1155.0	3.6	**0.1339**	0.008	5.8	ug/L	252	Standard
	Ni	60	2034.1	3.4	**0.9799**	0.027	2.8	ug/L	192	Standard
	Cu	65	13331.2	1.6	**6.9601**	0.242	3.5	ug/L	141	Standard
	Zn	66	18863.7	1.1	**16.4635**	0.229	1.4	ug/L	491	Standard
>	**Ge**	72	512762.7	2.0				ug/L	515087	Standard
	As	75	120.3	20.8	**0.1481**	0.025	16.6	ug/L	20	Standard
	Se	82	22.7	50.1	**0.0800**	0.103	128.5	ug/L	15	Standard
	Se-1	77	66.7	3.8	**0.1706**	0.046	27.2	ug/L	48	Standard
>	**Ga**	71	313.3	7.5				mg/L	90	Standard
	Rb	85	9062.7	6.6				ug/L	23	Standard
	Y	89	438095.5	1.1				ug/L	448510	Standard
>	**Rh**	103	11.7	65.5				ug/L	15	Standard
	Mo	98	749.7	5.2	**0.2622**	0.015	5.9	ug/L	11	Standard
	Ag	107	250.0	4.9	**0.0215**	0.002	9.3	ug/L	134	Standard
	Cd	111	268.7	5.4	**0.1246**	0.007	5.7	mg/L	7	Standard
	Cd	114	686.3	8.9	**0.1290**	0.012	9.4	ug/L	16	Standard
>	**In**	115	602735.5	0.4				ug/L	611117	Standard
	Sn	118	2001.8	9.6	**0.2412**	0.032	13.1	ug/L	518	Standard
	Sb	123	1551.7	1.3	**0.3206**	0.005	1.5	ug/L	127	Standard
	Ba	135	10171.4	0.7	**4.9702**	0.026	0.5	ug/L	45	Standard
	Ce	140	10203.4	3.4				ug/L	298	Standard
>	**Tb**	159	943711.5	1.3				ug/L	977942	Standard
	Ho	165	201.7	2.9				ug/L	12	Standard
	Tl	203	136.7	14.9	**0.0117**	0.003	21.7	ug/L	31	Standard
	Tl	205	126.7	4.6	**0.0153**	0.001	5.0	ug/L	20	Standard
	Pb	206	13666.8	0.5	**2.5119**	0.014	0.6	ug/L	286	Standard
	Pb	207	11351.3	1.7	**2.3133**	0.023	1.0	ug/L	257	Standard
	Pb	208	47777.6	0.4	**2.5197**	0.028	1.1	ug/L	956	Standard
	U	238	179.7	5.5	**0.0226**	0.001	5.2	ug/L	17	Standard
>	**Bi**	209	546981.3	0.8				ug/L	557080	Standard

Sample ID: L1606139302
Report Date/Time: Tuesday, June 28, 2016 22:36:40
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	296.7	10.2	**0.3250**	0.036	11.0	mg/L	32	Standard
K	39	131.7	25.3	**0.2905**	0.116	40.0	mg/L	30	Standard
Ca	43	28.3	44.4	**-1.1569**	1.919	165.9	mg/L	48	Standard
Fe	54	412.0	10.8	**0.2238**	0.038	16.8	mg/L	159	Standard
Fe	57	300.0	5.8	**0.3560**	0.058	16.2	mg/L	195	Standard
> **Sc-1**	45	40846.7	1.8				mg/L	39543	Standard
Cl	35	35351.2	2.0				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	2080.1	6.1				ug/L	1647	Standard
P	31	17501.8	0.9				ug/L	15231	Standard
S	34	2256.8	4.2				ug/L	1845	Standard
Sr	88	83.3	24.2				ug/L	70	Standard
C	12	216.7	41.9				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	30.0	66.7				mg/L	0	Standard
Dy	164	374.4	8.4				mg/L	12	Standard
Ho-1	165	201.7	2.9				mg/L	12	Standard
Er	166	186.7	17.2				mg/L	20	Standard
I	127	9588.0	8.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		93.551	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.549	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606139302
Report Date/Time: Tuesday, June 28, 2016 22:36:40
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.629
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.187
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139303
Sample Date/Time: Tuesday, June 28, 2016 22:37:34
Number of Replicates: 3
Autosampler Position: 345
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75224.7	2.8				ug/L	81760	Standard
	Be	9	26.7	47.2	**0.0195**	0.014	72.7	ug/L	17	Standard
	Al	27	335427.7	1.4	**2.3514**	0.039	1.7	ug/L	1410	Standard
	Sc	45	40873.4	3.2				ug/L	39543	Standard
	Ti	47	1306.7	3.3	**5.2014**	0.227	4.4	ug/L	39	Standard
	V	51	16739.3	1.3	**2.2527**	0.061	2.7	ug/L	1415	Standard
	Cr	52	12212.3	0.6	**0.7342**	0.019	2.6	ug/L	6946	Standard
	Cr	53	1208.4	9.2	**1.0637**	0.123	11.6	ug/L	328	Standard
	Mn	55	30356.4	1.8	**4.6331**	0.125	2.7	ug/L	823	Standard
	Co	59	971.0	0.7	**0.1098**	0.003	2.3	ug/L	252	Standard
	Ni	60	1423.4	2.9	**0.6666**	0.013	2.0	ug/L	192	Standard
	Cu	65	8494.4	2.1	**4.4744**	0.140	3.1	ug/L	141	Standard
	Zn	66	12099.5	0.6	**10.4446**	0.200	1.9	ug/L	491	Standard
>	**Ge**	72	504506.4	1.3				ug/L	515087	Standard
	As	75	95.4	18.1	**0.1269**	0.016	12.6	ug/L	20	Standard
	Se	82	22.1	16.3	**0.0773**	0.034	44.4	ug/L	15	Standard
	Se-1	77	59.3	9.3	**0.0847**	0.065	76.5	ug/L	48	Standard
>	**Ga**	71	286.7	13.3				mg/L	90	Standard
	Rb	85	4860.8	3.1				ug/L	23	Standard
	Y	89	447487.9	1.5				ug/L	448510	Standard
>	**Rh**	103	10.0	50.0				ug/L	15	Standard
	Mo	98	335.7	2.2	**0.1056**	0.003	3.1	ug/L	11	Standard
	Ag	107	149.3	7.8	**0.0061**	0.002	27.6	ug/L	134	Standard
	Cd	111	92.8	5.9	**0.0407**	0.003	6.7	mg/L	7	Standard
	Cd	114	275.4	18.0	**0.0483**	0.009	19.4	ug/L	16	Standard
>	**In**	115	608945.1	0.4				ug/L	611117	Standard
	Sn	118	2160.2	6.9	**0.2646**	0.024	9.0	ug/L	518	Standard
	Sb	123	1475.9	3.1	**0.3011**	0.011	3.5	ug/L	127	Standard
	Ba	135	7298.8	1.0	**3.5230**	0.049	1.4	ug/L	45	Standard
	Ce	140	7186.7	3.0				ug/L	298	Standard
>	**Tb**	159	941207.0	0.6				ug/L	977942	Standard
	Ho	165	175.0	29.8				ug/L	12	Standard
	Tl	203	127.3	10.7	**0.0108**	0.002	16.0	ug/L	31	Standard
	Tl	205	91.7	22.0	**0.0106**	0.003	26.7	ug/L	20	Standard
	Pb	206	10382.9	0.6	**1.9175**	0.012	0.6	ug/L	286	Standard
	Pb	207	8724.8	0.9	**1.7876**	0.028	1.6	ug/L	257	Standard
	Pb	208	36599.7	1.1	**1.9407**	0.035	1.8	ug/L	956	Standard
	U	238	138.7	9.6	**0.0170**	0.002	11.7	ug/L	17	Standard
>	**Bi**	209	540229.2	0.6				ug/L	557080	Standard

Sample ID: L1606139303
Report Date/Time: Tuesday, June 28, 2016 22:39:51
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	201.7	23.0	**0.2027**	0.061	30.0	mg/L	32	Standard
K	39	85.0	15.6	**0.1365**	0.043	31.8	mg/L	30	Standard
Ca	43	46.7	34.4	**1.4659**	2.107	143.8	mg/L	48	Standard
Fe	54	348.6	7.5	**0.1680**	0.018	11.0	mg/L	159	Standard
Fe	57	325.0	9.4	**0.4364**	0.086	19.7	mg/L	195	Standard
> **Sc-1**	45	40873.4	3.2				mg/L	39543	Standard
Cl	35	35045.8	0.8				ug/L	27642	Standard
Kr	83	2.3	107.9				ug/L	2	Standard
Br	81	1873.4	9.4				ug/L	1647	Standard
P	31	17319.9	1.5				ug/L	15231	Standard
S	34	2126.8	6.5				ug/L	1845	Standard
Sr	88	78.3	28.8				ug/L	70	Standard
C	12	160.0	25.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	78.1				mg/L	0	Standard
Dy	164	223.9	25.3				mg/L	12	Standard
Ho-1	165	175.0	29.8				mg/L	12	Standard
Er	166	126.7	25.4				mg/L	20	Standard
I	127	7363.5	5.1				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		92.007	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		97.946	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606139303
Report Date/Time: Tuesday, June 28, 2016 22:39:51
Page 2

Approved: June 29, 2016

⌊	Rb	85	
⌈	Y	89	
⌊>	Rh	103	
⌈	Mo	98	
|	Ag	107	
|	Cd	111	
|	Cd	114	
|>	In	115	99.645
|	Sn	118	
|	Sb	123	
⌊	Ba	135	
⌈	Ce	140	
⌊>	Tb	159	
⌈	Ho	165	
|	Tl	203	
|	Tl	205	
|	Pb	206	
|	Pb	207	
|	Pb	208	
|	U	238	
⌊>	Bi	209	96.975
⌈	Na	23	
|	Mg	24	
|	K	39	
|	Ca	43	
|	Fe	54	
|	Fe	57	
⌊>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139304
Sample Date/Time: Tuesday, June 28, 2016 22:40:46
Number of Replicates: 3
Autosampler Position: 346
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74628.4	2.8				ug/L	81760	Standard
	Be	9	20.0	25.0	**0.0122**	0.005	40.5	ug/L	17	Standard
	Al	27	327323.4	5.6	**2.3109**	0.073	3.1	ug/L	1410	Standard
	Sc	45	39232.4	1.5				ug/L	39543	Standard
	Ti	47	1081.7	2.8	**4.2827**	0.197	4.6	ug/L	39	Standard
	V	51	17364.8	3.6	**2.3462**	0.053	2.3	ug/L	1415	Standard
	Cr	52	11969.4	3.2	**0.6998**	0.018	2.6	ug/L	6946	Standard
	Cr	53	1151.7	1.8	**0.9999**	0.045	4.5	ug/L	328	Standard
	Mn	55	27710.7	5.0	**4.2196**	0.194	4.6	ug/L	823	Standard
	Co	59	805.7	4.4	**0.0858**	0.004	4.6	ug/L	252	Standard
	Ni	60	1156.4	2.7	**0.5226**	0.014	2.7	ug/L	192	Standard
	Cu	65	7318.8	3.6	**3.8448**	0.072	1.9	ug/L	141	Standard
	Zn	66	9717.1	4.5	**8.2301**	0.259	3.2	ug/L	491	Standard
>	**Ge**	72	504100.5	3.0				ug/L	515087	Standard
	As	75	86.2	19.2	**0.1185**	0.013	11.2	ug/L	20	Standard
	Se	82	21.0	26.2	**0.0675**	0.052	77.2	ug/L	15	Standard
	Se-1	77	57.7	9.6	**0.0648**	0.094	144.5	ug/L	48	Standard
>	**Ga**	71	233.3	13.9				mg/L	90	Standard
	Rb	85	3882.2	6.3				ug/L	23	Standard
	Y	89	430223.1	5.6				ug/L	448510	Standard
>	**Rh**	103	8.3	124.9				ug/L	15	Standard
	Mo	98	801.5	6.3	**0.2860**	0.012	4.3	ug/L	11	Standard
	Ag	107	120.7	5.8	**0.0023**	0.000	18.3	ug/L	134	Standard
	Cd	111	92.6	9.2	**0.0417**	0.004	9.3	mg/L	7	Standard
	Cd	114	241.3	24.7	**0.0431**	0.013	29.2	ug/L	16	Standard
>	**In**	115	593856.9	3.7				ug/L	611117	Standard
	Sn	118	1755.1	4.0	**0.2036**	0.011	5.3	ug/L	518	Standard
	Sb	123	1169.8	3.7	**0.2423**	0.004	1.7	ug/L	127	Standard
	Ba	135	5956.2	4.1	**2.9440**	0.065	2.2	ug/L	45	Standard
	Ce	140	5319.3	7.3				ug/L	298	Standard
>	**Tb**	159	910819.3	2.0				ug/L	977942	Standard
	Ho	165	128.3	12.5				ug/L	12	Standard
	Tl	203	88.0	5.7	**0.0062**	0.001	12.2	ug/L	31	Standard
	Tl	205	70.0	25.8	**0.0077**	0.003	33.5	ug/L	20	Standard
	Pb	206	8718.5	3.6	**1.6271**	0.072	4.4	ug/L	286	Standard
	Pb	207	7163.0	4.3	**1.4818**	0.072	4.9	ug/L	257	Standard
	Pb	208	29853.9	4.1	**1.5981**	0.065	4.1	ug/L	956	Standard
	U	238	115.3	2.6	**0.0139**	0.000	1.6	ug/L	17	Standard
>	**Bi**	209	531672.0	1.7				ug/L	557080	Standard

Sample ID: L1606139304
Report Date/Time: Tuesday, June 28, 2016 22:43:03
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	220.0	2.3	**0.2381**	0.006	2.4	mg/L	32	Standard
K	39	73.3	3.9	**0.1083**	0.007	6.4	mg/L	30	Standard
Ca	43	31.7	24.1	**-0.5049**	1.115	220.9	mg/L	48	Standard
Fe	54	263.9	12.4	**0.1035**	0.028	27.1	mg/L	159	Standard
Fe	57	248.3	16.1	**0.2213**	0.145	65.6	mg/L	195	Standard
> **Sc-1**	45	39232.4	1.5				mg/L	39543	Standard
Cl	35	34033.6	5.9				ug/L	27642	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	1946.8	12.8				ug/L	1647	Standard
P	31	16857.7	2.7				ug/L	15231	Standard
S	34	2100.1	4.5				ug/L	1845	Standard
Sr	88	66.7	28.4				ug/L	70	Standard
C	12	176.7	23.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0					mg/L	0	Standard
Dy	164	134.9	39.4				mg/L	12	Standard
Ho-1	165	128.3	12.5				mg/L	12	Standard
Er	166	106.7	10.8				mg/L	20	Standard
I	127	6791.5	2.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.277	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.867	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606139304
Report Date/Time: Tuesday, June 28, 2016 22:43:03
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.176
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.439
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139305
Sample Date/Time: Tuesday, June 28, 2016 22:43:57
Number of Replicates: 3
Autosampler Position: 347
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74899.7	0.9				ug/L	81760	Standard
	Be	9	18.3	15.7	**0.0104**	0.003	29.0	ug/L	17	Standard
	Al	27	328437.1	1.5	**2.3115**	0.035	1.5	ug/L	1410	Standard
	Sc	45	40647.8	2.1				ug/L	39543	Standard
	Ti	47	1117.7	1.6	**4.2600**	0.050	1.2	ug/L	39	Standard
	V	51	13332.6	0.8	**1.6821**	0.023	1.4	ug/L	1415	Standard
	Cr	52	11640.1	0.5	**0.5893**	0.015	2.6	ug/L	6946	Standard
	Cr	53	1158.4	10.6	**0.9575**	0.132	13.8	ug/L	328	Standard
	Mn	55	25508.1	0.7	**3.7255**	0.023	0.6	ug/L	823	Standard
	Co	59	795.0	2.3	**0.0800**	0.002	2.0	ug/L	252	Standard
	Ni	60	1111.4	4.5	**0.4759**	0.022	4.7	ug/L	192	Standard
	Cu	65	7781.0	2.0	**3.9419**	0.127	3.2	ug/L	141	Standard
	Zn	66	11686.8	2.5	**9.6690**	0.148	1.5	ug/L	491	Standard
>	**Ge**	72	523177.1	1.2				ug/L	515087	Standard
	As	75	150.7	20.5	**0.1724**	0.029	16.5	ug/L	20	Standard
	Se	82	42.9	7.7	**0.2478**	0.029	11.9	ug/L	15	Standard
	Se-1	77	60.7	2.5	**0.0739**	0.016	21.3	ug/L	48	Standard
>	**Ga**	71	226.7	1.3				mg/L	90	Standard
	Rb	85	5315.9	7.9				ug/L	23	Standard
	Y	89	431953.8	1.9				ug/L	448510	Standard
>	**Rh**	103	15.0	66.7				ug/L	15	Standard
	Mo	98	276.3	6.3	**0.0857**	0.006	6.9	ug/L	11	Standard
	Ag	107	169.0	7.2	**0.0096**	0.002	19.7	ug/L	134	Standard
	Cd	111	162.4	6.5	**0.0750**	0.005	6.0	mg/L	7	Standard
	Cd	114	335.5	8.9	**0.0613**	0.005	8.9	ug/L	16	Standard
>	**In**	115	595959.3	0.7				ug/L	611117	Standard
	Sn	118	1471.7	5.0	**0.1534**	0.014	9.4	ug/L	518	Standard
	Sb	123	1173.4	5.5	**0.2422**	0.014	5.7	ug/L	127	Standard
	Ba	135	5837.1	3.3	**2.8740**	0.084	2.9	ug/L	45	Standard
	Ce	140	5496.0	2.0				ug/L	298	Standard
>	**Tb**	159	937149.5	1.0				ug/L	977942	Standard
	Ho	165	138.3	7.5				ug/L	12	Standard
	Tl	203	94.3	16.1	**0.0066**	0.002	26.5	ug/L	31	Standard
	Tl	205	66.7	26.3	**0.0070**	0.002	32.8	ug/L	20	Standard
	Pb	206	8082.5	0.7	**1.4592**	0.036	2.4	ug/L	286	Standard
	Pb	207	6688.5	1.5	**1.3390**	0.041	3.1	ug/L	257	Standard
	Pb	208	27392.7	1.4	**1.4181**	0.004	0.2	ug/L	956	Standard
	U	238	108.7	8.4	**0.0125**	0.002	12.3	ug/L	17	Standard
>	**Bi**	209	547233.0	1.6				ug/L	557080	Standard

Sample ID: L1606139305
Report Date/Time: Tuesday, June 28, 2016 22:46:14
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.3347**	17.892	173.1	mg/L	0	Standard
Mg	24	221.7	15.4	**0.2295**	0.039	17.1	mg/L	32	Standard
K	39	100.0	17.3	**0.1871**	0.055	29.3	mg/L	30	Standard
Ca	43	26.7	60.3	**-1.4304**	2.300	160.8	mg/L	48	Standard
Fe	54	289.4	9.6	**0.1178**	0.025	21.2	mg/L	159	Standard
Fe	57	271.7	9.3	**0.2686**	0.097	36.1	mg/L	195	Standard
Sc-1	45	40647.8	2.1				mg/L	39543	Standard
Cl	35	33690.1	1.5				ug/L	27642	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	7261.7	1.9				ug/L	1647	Standard
P	31	16325.5	0.4				ug/L	15231	Standard
S	34	1950.1	10.4				ug/L	1845	Standard
Sr	88	81.7	9.4				ug/L	70	Standard
C	12	186.7	18.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	222.3	16.3				mg/L	12	Standard
Ho-1	165	138.3	7.5				mg/L	12	Standard
Er	166	160.0	6.3				mg/L	20	Standard
I	127	6493.1	4.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.609	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.571	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606139305
Report Date/Time: Tuesday, June 28, 2016 22:46:14
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.520
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.232
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 22:47:10
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	71793.2	3.7				ug/L	81760	Standard
	Be	9	44618.9	1.2	**51.1958**	1.696	3.3	ug/L	17	Standard
	Al	27	6865175.8	2.0	**50.7482**	1.963	3.9	ug/L	1410	Standard
	Sc	45	39468.0	3.1				ug/L	39543	Standard
	Ti	47	23906.5	1.3	**96.7870**	2.402	2.5	ug/L	39	Standard
	V	51	334011.6	0.6	**48.2557**	0.941	2.0	ug/L	1415	Standard
	Cr	52	340377.2	0.5	**48.2561**	1.001	2.1	ug/L	6946	Standard
	Cr	53	42902.3	1.0	**48.9484**	1.164	2.4	ug/L	328	Standard
	Mn	55	309303.1	1.2	**47.9606**	0.591	1.2	ug/L	823	Standard
	Co	59	332037.9	0.8	**47.6215**	0.384	0.8	ug/L	252	Standard
	Ni	60	90431.4	0.3	**48.3175**	0.794	1.6	ug/L	192	Standard
	Cu	65	91394.5	0.3	**48.4155**	0.851	1.8	ug/L	141	Standard
	Zn	66	53482.9	1.3	**48.3851**	1.341	2.8	ug/L	491	Standard
>	**Ge**	72	510859.4	1.5				ug/L	515087	Standard
	As	75	54007.5	0.9	**48.4846**	1.105	2.3	ug/L	20	Standard
	Se	82	5360.1	0.8	**46.8280**	0.601	1.3	ug/L	15	Standard
	Se-1	77	3686.4	1.7	**48.6230**	0.812	1.7	ug/L	48	Standard
>	**Ga**	71	56.7	18.4				mg/L	90	Standard
	Rb	85	810.0	4.5				ug/L	23	Standard
	Y	89	431931.2	2.6				ug/L	448510	Standard
>	**Rh**	103	26.7	43.3				ug/L	15	Standard
	Mo	98	251225.1	0.7	**96.8468**	1.378	1.4	ug/L	11	Standard
	Ag	107	304440.8	0.4	**47.2505**	0.474	1.0	ug/L	134	Standard
	Cd	111	96060.5	0.9	**46.8341**	0.834	1.8	mg/L	7	Standard
	Cd	114	238889.8	1.2	**47.9103**	0.301	0.6	ug/L	16	Standard
>	**In**	115	586435.3	1.3				ug/L	611117	Standard
	Sn	118	269259.0	2.4	**47.3660**	1.700	3.6	ug/L	518	Standard
	Sb	123	219432.6	0.9	**48.4541**	0.360	0.7	ug/L	127	Standard
	Ba	135	92563.9	0.7	**46.7027**	0.660	1.4	ug/L	45	Standard
	Ce	140	88.3	61.1				ug/L	298	Standard
>	**Tb**	159	918919.9	0.8				ug/L	977942	Standard
	Ho	165	13.3	21.7				ug/L	12	Standard
	Tl	203	386832.4	0.9	**48.5637**	0.777	1.6	ug/L	31	Standard
	Tl	205	343277.1	1.0	**49.6247**	0.132	0.3	ug/L	20	Standard
	Pb	206	249448.7	0.4	**48.7606**	0.291	0.6	ug/L	286	Standard
	Pb	207	228146.2	0.6	**49.4471**	0.531	1.1	ug/L	257	Standard
	Pb	208	902842.6	0.2	**50.5774**	0.397	0.8	ug/L	956	Standard
	U	238	333363.3	0.9	**49.2162**	0.753	1.5	ug/L	17	Standard
>	**Bi**	209	526646.5	0.9				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:49:27
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.4874**	18.156	173.1	mg/L	0	Standard
Mg	24	3567.1	3.5	**4.7067**	0.317	6.7	mg/L	32	Standard
K	39	1421.7	5.7	**4.6773**	0.373	8.0	mg/L	30	Standard
Ca	43	60.0	8.3	**3.7779**	0.573	15.2	mg/L	48	Standard
Fe	54	5303.1	4.9	**4.6695**	0.087	1.9	mg/L	159	Standard
Fe	57	1570.1	3.9	**4.6916**	0.328	7.0	mg/L	195	Standard
> **Sc-1**	45	39468.0	3.1				mg/L	39543	Standard
Cl	35	34178.5	1.9				ug/L	27642	Standard
Kr	83	1.7	124.9				ug/L	2	Standard
Br	81	1696.8	1.5				ug/L	1647	Standard
P	31	17313.2	2.3				ug/L	15231	Standard
S	34	2535.2	5.3				ug/L	1845	Standard
Sr	88	81.7	40.8				ug/L	70	Standard
C	12	120.0	14.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	8.9	11.2				mg/L	12	Standard
Ho-1	165	13.3	21.7				mg/L	12	Standard
Er	166	23.3	89.2				mg/L	20	Standard
I	127	2468.5	18.6				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	102.392		
	Al	27	101.496		
	Sc	45			
	Ti	47	96.787		
	V	51	96.511		
	Cr	52	96.512		
	Cr	53			
	Mn	55	95.921		
	Co	59	95.243		
	Ni	60	96.635		
	Cu	65	96.831		
	Zn	66	96.770		
>	Ge	72		99.179	
	As	75	96.969		
	Se	82	93.656		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 22:49:27
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	96.847		
Ag	107	94.501		
Cd	111	93.668		
Cd	114			
> In	115			95.961
Sn	118	94.732		
Sb	123	96.908		
Ba	135	93.405		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	97.127		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.155		
U	238	98.432		
> Bi	209			94.537
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 22:50:22
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72519.8	1.9				ug/L	81760	Standard
	Be	9	21.7	113.8	**0.0146**	0.027	188.1	ug/L	17	Standard
	Al	27	2256.8	25.0	**0.0021**	0.004	187.1	ug/L	1410	Standard
	Sc	45	39016.8	1.3				ug/L	39543	Standard
	Ti	47	28.7	19.2	**-0.0470**	0.020	43.5	ug/L	39	Standard
	V	51	1194.9	2.1	**-0.0301**	0.006	21.3	ug/L	1415	Standard
	Cr	52	6354.0	0.6	**-0.1199**	0.015	12.7	ug/L	6946	Standard
	Cr	53	495.0	4.6	**0.2373**	0.020	8.3	ug/L	328	Standard
	Mn	55	724.7	7.6	**-0.0338**	0.007	19.6	ug/L	823	Standard
	Co	59	253.7	9.1	**0.0057**	0.003	47.4	ug/L	252	Standard
	Ni	60	174.7	5.0	**-0.0098**	0.003	33.3	ug/L	192	Standard
	Cu	65	124.0	10.1	**-0.0239**	0.006	23.4	ug/L	141	Standard
	Zn	66	416.7	1.4	**-0.4373**	0.011	2.5	ug/L	491	Standard
>	**Ge**	72	501758.9	1.8				ug/L	515087	Standard
	As	75	-30.1	127.3	**0.0123**	0.035	285.9	ug/L	20	Standard
	Se	82	17.1	40.3	**0.0328**	0.058	178.2	ug/L	15	Standard
	Se-1	77	51.3	18.1	**-0.0207**	0.113	545.7	ug/L	48	Standard
>	**Ga**	71	38.3	27.2				mg/L	90	Standard
	Rb	85	28.3	36.7				ug/L	23	Standard
	Y	89	432168.4	0.3				ug/L	448510	Standard
>	**Rh**	103	15.0	66.7				ug/L	15	Standard
	Mo	98	274.3	30.4	**0.0862**	0.033	37.9	ug/L	11	Standard
	Ag	107	126.3	11.5	**0.0033**	0.002	71.2	ug/L	134	Standard
	Cd	111	15.5	45.6	**0.0046**	0.003	75.0	mg/L	7	Standard
	Cd	114	32.8	69.1	**0.0016**	0.005	280.8	ug/L	16	Standard
>	**In**	115	589750.3	1.0				ug/L	611117	Standard
	Sn	118	705.0	18.1	**0.0215**	0.021	98.3	ug/L	518	Standard
	Sb	123	430.7	24.1	**0.0816**	0.022	27.0	ug/L	127	Standard
	Ba	135	35.7	21.2	**-0.0072**	0.004	55.2	ug/L	45	Standard
	Ce	140	38.3	41.9				ug/L	298	Standard
>	**Tb**	159	920829.0	0.8				ug/L	977942	Standard
	Ho	165	5.0	100.0				ug/L	12	Standard
	Tl	203	59.7	35.8	**0.0027**	0.003	98.6	ug/L	31	Standard
	Tl	205	48.3	29.9	**0.0047**	0.002	43.8	ug/L	20	Standard
	Pb	206	270.0	6.9	**-0.0110**	0.004	33.8	ug/L	286	Standard
	Pb	207	257.3	3.4	**-0.0024**	0.002	65.9	ug/L	257	Standard
	Pb	208	1000.7	8.0	**-0.0046**	0.004	92.2	ug/L	956	Standard
	U	238	61.0	23.2	**0.0060**	0.002	34.0	ug/L	17	Standard
>	**Bi**	209	528739.4	0.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:52:39
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	23.3	86.6	**-0.0259**	0.027	103.6	mg/L	32	Standard
K	39	31.7	48.2	**-0.0333**	0.052	155.8	mg/L	30	Standard
Ca	43	30.0	72.6	**-0.7014**	3.391	483.5	mg/L	48	Standard
Fe	54	129.4	20.1	**-0.0182**	0.025	139.3	mg/L	159	Standard
Fe	57	250.0	16.4	**0.2316**	0.148	64.1	mg/L	195	Standard
> **Sc-1**	45	39016.8	1.3				mg/L	39543	Standard
Cl	35	33625.3	1.5				ug/L	27642	Standard
Kr	83	3.3	124.9				ug/L	2	Standard
Br	81	1580.1	13.5				ug/L	1647	Standard
P	31	17735.4	1.8				ug/L	15231	Standard
S	34	2481.9	5.6				ug/L	1845	Standard
Sr	88	73.3	14.2				ug/L	70	Standard
C	12	113.3	41.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	0	Standard
Dy	164	9.4	5.9				mg/L	12	Standard
Ho-1	165	5.0	100.0				mg/L	12	Standard
Er	166	13.3	86.6				mg/L	20	Standard
I	127	1860.1	15.9				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		97.412	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:52:39
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.504
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.913
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 22:52:39
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139306
Sample Date/Time: Tuesday, June 28, 2016 22:53:35
Number of Replicates: 3
Autosampler Position: 348
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70220.8	3.4				ug/L	81760	Standard
	Be	9	21.7	48.0	**0.0156**	0.012	75.7	ug/L	17	Standard
	Al	27	339019.6	1.3	**2.5478**	0.066	2.6	ug/L	1410	Standard
	Sc	45	39077.0	1.3				ug/L	39543	Standard
	Ti	47	1359.1	3.4	**5.3458**	0.211	3.9	ug/L	39	Standard
	V	51	13893.8	2.6	**1.8093**	0.072	4.0	ug/L	1415	Standard
	Cr	52	10697.1	2.4	**0.4928**	0.058	11.9	ug/L	6946	Standard
	Cr	53	1106.7	6.2	**0.9306**	0.098	10.5	ug/L	328	Standard
	Mn	55	27827.9	2.7	**4.1809**	0.179	4.3	ug/L	823	Standard
	Co	59	887.0	0.6	**0.0960**	0.003	2.8	ug/L	252	Standard
	Ni	60	1041.0	6.1	**0.4528**	0.043	9.5	ug/L	192	Standard
	Cu	65	6718.2	2.4	**3.4752**	0.137	3.9	ug/L	141	Standard
	Zn	66	10492.3	2.9	**8.8264**	0.376	4.3	ug/L	491	Standard
>	**Ge**	72	510902.9	1.6				ug/L	515087	Standard
	As	75	95.4	28.2	**0.1256**	0.023	18.4	ug/L	20	Standard
	Se	82	17.2	18.1	**0.0307**	0.025	82.4	ug/L	15	Standard
	Se-1	77	52.3	18.1	**-0.0171**	0.139	811.5	ug/L	48	Standard
>	**Ga**	71	256.7	6.0				mg/L	90	Standard
	Rb	85	4037.2	4.0				ug/L	23	Standard
	Y	89	441838.7	1.2				ug/L	448510	Standard
>	**Rh**	103	13.3	43.3				ug/L	15	Standard
	Mo	98	321.3	2.1	**0.1057**	0.004	4.2	ug/L	11	Standard
	Ag	107	134.7	13.7	**0.0048**	0.003	55.1	ug/L	134	Standard
	Cd	111	94.4	13.5	**0.0434**	0.006	14.4	mg/L	7	Standard
	Cd	114	184.5	30.3	**0.0322**	0.011	33.4	ug/L	16	Standard
>	**In**	115	582517.6	1.6				ug/L	611117	Standard
	Sn	118	1633.4	8.5	**0.1881**	0.027	14.5	ug/L	518	Standard
	Sb	123	1202.8	2.0	**0.2547**	0.009	3.5	ug/L	127	Standard
	Ba	135	5625.0	0.7	**2.8338**	0.047	1.7	ug/L	45	Standard
	Ce	140	7943.7	1.7				ug/L	298	Standard
>	**Tb**	159	895525.5	0.4				ug/L	977942	Standard
	Ho	165	181.7	15.7				ug/L	12	Standard
	Tl	203	104.3	12.5	**0.0084**	0.002	20.6	ug/L	31	Standard
	Tl	205	91.7	22.0	**0.0111**	0.003	27.4	ug/L	20	Standard
	Pb	206	6727.8	0.5	**1.2603**	0.018	1.4	ug/L	286	Standard
	Pb	207	5614.0	0.6	**1.1666**	0.017	1.4	ug/L	257	Standard
	Pb	208	23664.6	2.1	**1.2740**	0.038	3.0	ug/L	956	Standard
	U	238	141.7	17.4	**0.0181**	0.004	20.7	ug/L	17	Standard
>	**Bi**	209	523874.2	0.9				ug/L	557080	Standard

Sample ID: L1606139306
Report Date/Time: Tuesday, June 28, 2016 22:55:52
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	196.7	25.5	**0.2078**	0.067	32.2	mg/L	32	Standard
K	39	58.3	32.5	**0.0586**	0.068	115.6	mg/L	30	Standard
Ca	43	33.3	48.2	**-0.2427**	2.425	998.9	mg/L	48	Standard
Fe	54	342.2	5.9	**0.1765**	0.023	12.8	mg/L	159	Standard
Fe	57	293.3	7.1	**0.3778**	0.076	20.2	mg/L	195	Standard
Sc-1	45	39077.0	1.3				mg/L	39543	Standard
Cl	35	33443.5	2.1				ug/L	27642	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	1690.1	11.7				ug/L	1647	Standard
P	31	17176.4	4.2				ug/L	15231	Standard
S	34	2306.8	4.6				ug/L	1845	Standard
Sr	88	88.3	31.2				ug/L	70	Standard
C	12	143.3	10.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	86.6				mg/L	0	Standard
Dy	164	280.6	4.7				mg/L	12	Standard
Ho-1	165	181.7	15.7				mg/L	12	Standard
Er	166	196.7	36.8				mg/L	20	Standard
I	127	4592.4	1.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		85.886	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.188	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606139306
Report Date/Time: Tuesday, June 28, 2016 22:55:52
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.320
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.039
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606139307
Sample Date/Time: Tuesday, June 28, 2016 22:56:46
Number of Replicates: 3
Autosampler Position: 349
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72494.7	0.8				ug/L	81760	Standard
	Be	9	18.3	15.7	**0.0111**	0.003	30.5	ug/L	17	Standard
	Al	27	342418.0	1.1	**2.4911**	0.036	1.4	ug/L	1410	Standard
	Sc	45	40101.3	1.0				ug/L	39543	Standard
	Ti	47	932.7	2.7	**3.5634**	0.149	4.2	ug/L	39	Standard
	V	51	14612.8	2.1	**1.8826**	0.024	1.3	ug/L	1415	Standard
	Cr	52	11386.6	1.0	**0.5692**	0.057	10.1	ug/L	6946	Standard
	Cr	53	1260.1	6.9	**1.0869**	0.132	12.1	ug/L	328	Standard
	Mn	55	23640.1	1.0	**3.4768**	0.122	3.5	ug/L	823	Standard
	Co	59	741.4	0.2	**0.0735**	0.003	3.5	ug/L	252	Standard
	Ni	60	1110.0	3.0	**0.4811**	0.030	6.3	ug/L	192	Standard
	Cu	65	7534.9	0.8	**3.8515**	0.130	3.4	ug/L	141	Standard
	Zn	66	11297.5	1.4	**9.4161**	0.135	1.4	ug/L	491	Standard
>	**Ge**	72	518404.2	2.6				ug/L	515087	Standard
	As	75	185.7	13.2	**0.2041**	0.018	8.6	ug/L	20	Standard
	Se	82	58.4	11.2	**0.3845**	0.051	13.1	ug/L	15	Standard
	Se-1	77	57.0	12.3	**0.0333**	0.095	284.1	ug/L	48	Standard
>	**Ga**	71	201.7	7.6				mg/L	90	Standard
	Rb	85	5255.9	2.3				ug/L	23	Standard
	Y	89	429044.9	0.4				ug/L	448510	Standard
>	**Rh**	103	18.3	68.6				ug/L	15	Standard
	Mo	98	299.7	1.8	**0.0961**	0.002	1.9	ug/L	11	Standard
	Ag	107	153.3	22.0	**0.0075**	0.005	67.1	ug/L	134	Standard
	Cd	111	143.0	17.2	**0.0665**	0.011	16.8	mg/L	7	Standard
	Cd	114	356.8	31.1	**0.0663**	0.022	32.5	ug/L	16	Standard
>	**In**	115	588246.1	1.2				ug/L	611117	Standard
	Sn	118	1550.1	10.3	**0.1702**	0.025	14.7	ug/L	518	Standard
	Sb	123	1125.9	6.3	**0.2350**	0.013	5.6	ug/L	127	Standard
	Ba	135	5627.4	4.0	**2.8066**	0.107	3.8	ug/L	45	Standard
	Ce	140	5279.3	1.4				ug/L	298	Standard
>	**Tb**	159	929768.7	1.5				ug/L	977942	Standard
	Ho	165	90.0	24.2				ug/L	12	Standard
	Tl	203	125.7	60.8	**0.0104**	0.009	86.4	ug/L	31	Standard
	Tl	205	76.7	65.3	**0.0084**	0.007	82.2	ug/L	20	Standard
	Pb	206	9321.2	0.6	**1.7003**	0.038	2.2	ug/L	286	Standard
	Pb	207	7700.6	1.2	**1.5576**	0.043	2.8	ug/L	257	Standard
	Pb	208	32126.4	0.5	**1.6815**	0.027	1.6	ug/L	956	Standard
	U	238	95.0	19.2	**0.0106**	0.002	23.3	ug/L	17	Standard
>	**Bi**	209	544815.4	1.6				ug/L	557080	Standard

Sample ID: L1606139307
Report Date/Time: Tuesday, June 28, 2016 22:59:03
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	241.7	8.4	**0.2600**	0.024	9.1	mg/L	32	Standard
K	39	93.3	13.5	**0.1697**	0.044	25.8	mg/L	30	Standard
Ca	43	48.3	11.9	**1.8938**	0.874	46.1	mg/L	48	Standard
Fe	54	296.6	3.5	**0.1277**	0.009	7.4	mg/L	159	Standard
Fe	57	275.0	13.7	**0.2917**	0.134	46.0	mg/L	195	Standard
> **Sc-1**	45	40101.3	1.0				mg/L	39543	Standard
Cl	35	33371.4	1.5				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	12191.6	6.9				ug/L	1647	Standard
P	31	16642.5	2.2				ug/L	15231	Standard
S	34	2221.8	1.3				ug/L	1845	Standard
Sr	88	81.7	3.5				ug/L	70	Standard
C	12	180.0	14.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	78.1				mg/L	0	Standard
Dy	164	179.2	25.7				mg/L	12	Standard
Ho-1	165	90.0	24.2				mg/L	12	Standard
Er	166	86.7	26.6				mg/L	20	Standard
I	127	5902.8	5.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		88.668	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.644	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606139307
Report Date/Time: Tuesday, June 28, 2016 22:59:03
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.258
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.798
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606140101
Sample Date/Time: Tuesday, June 28, 2016 22:59:58
Number of Replicates: 3
Autosampler Position: 350
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81815.4	2.6				ug/L	81760	Standard
	Be	9	21.7	13.3	**0.0121**	0.003	22.9	ug/L	17	Standard
	Al	27	45785308.7	4.2	**296.7336**	5.244	1.8	ug/L	1410	Standard
	Sc	45	41065.6	3.8				ug/L	39543	Standard
	Ti	47	178.0	12.2	**0.5887**	0.089	15.1	ug/L	39	Standard
	V	51	35448.4	4.1	**5.1643**	0.153	3.0	ug/L	1415	Standard
	Cr	52	9787.8	5.7	**0.4235**	0.068	16.1	ug/L	6946	Standard
	Cr	53	17411.8	14.2	**20.5297**	2.734	13.3	ug/L	328	Standard
	Mn	55	165972.1	5.7	**26.8086**	1.220	4.6	ug/L	823	Standard
	Co	59	1056.4	5.8	**0.1270**	0.007	5.7	ug/L	252	Standard
	Ni	60	3143.3	7.3	**1.6522**	0.109	6.6	ug/L	192	Standard
	Cu	65	2057.8	4.7	**1.0496**	0.040	3.9	ug/L	141	Standard
	Zn	66	6024.5	5.7	**4.9599**	0.261	5.3	ug/L	491	Standard
>	**Ge**	72	489001.5	1.3				ug/L	515087	Standard
	As	75	1112.2	8.9	**1.0817**	0.080	7.4	ug/L	20	Standard
	Se	82	302.7	9.2	**2.6482**	0.222	8.4	ug/L	15	Standard
	Se-1	77	1245.1	18.8	**16.6637**	3.078	18.5	ug/L	48	Standard
>	**Ga**	71	58.3	55.1				mg/L	90	Standard
	Rb	85	2691.9	5.7				ug/L	23	Standard
	Y	89	432812.2	6.0				ug/L	448510	Standard
>	**Rh**	103	128.3	14.8				ug/L	15	Standard
	Mo	98	642.1	7.2	**0.2310**	0.011	4.8	ug/L	11	Standard
	Ag	107	89.7	20.6	**-0.0021**	0.003	162.7	ug/L	134	Standard
	Cd	111	183.2	13.1	**0.0876**	0.014	15.6	mg/L	7	Standard
	Cd	114	423.4	6.2	**0.0810**	0.006	6.9	ug/L	16	Standard
>	**In**	115	580206.4	3.0				ug/L	611117	Standard
	Sn	118	1115.0	4.7	**0.0968**	0.010	9.8	ug/L	518	Standard
	Sb	123	200.7	10.2	**0.0319**	0.003	10.9	ug/L	127	Standard
	Ba	135	44440.1	3.3	**22.6462**	0.082	0.4	ug/L	45	Standard
	Ce	140	446.7	12.1				ug/L	298	Standard
>	**Tb**	159	929235.2	2.2				ug/L	977942	Standard
	Ho	165	38.3	7.5				ug/L	12	Standard
	Tl	203	126.0	16.1	**0.0119**	0.003	21.7	ug/L	31	Standard
	Tl	205	103.3	17.0	**0.0135**	0.003	21.8	ug/L	20	Standard
	Pb	206	396.7	5.3	**0.0183**	0.002	11.9	ug/L	286	Standard
	Pb	207	341.7	4.5	**0.0202**	0.002	8.5	ug/L	257	Standard
	Pb	208	1505.0	2.0	**0.0286**	0.002	5.8	ug/L	956	Standard
	U	238	10687.1	4.8	**1.6615**	0.054	3.3	ug/L	17	Standard
>	**Bi**	209	499114.9	2.7				ug/L	557080	Standard

Sample ID: L1606140101
Report Date/Time: Tuesday, June 28, 2016 23:02:15
Page 1

Approved: June 29, 2016

Na	23	8.3	91.7	**52.4138**	47.655	90.9	mg/L	0	Standard
Mg	24	165387.3	4.5	**211.9511**	6.406	3.0	mg/L	32	Standard
K	39	600.0	13.6	**1.8174**	0.318	17.5	mg/L	30	Standard
Ca	43	581.7	9.4	**79.6243**	5.149	6.5	mg/L	48	Standard
Fe	54	126.1	20.1	**-0.0269**	0.024	89.8	mg/L	159	Standard
Fe	57	591.7	6.9	**1.2999**	0.139	10.7	mg/L	195	Standard
> **Sc-1**	45	41065.6	3.8				mg/L	39543	Standard
Cl	35	47499.7	3.6				ug/L	27642	Standard
Kr	83	0.7	173.2				ug/L	2	Standard
Br	81	60832.3	7.6				ug/L	1647	Standard
P	31	21663.8	6.1				ug/L	15231	Standard
S	34	2320.2	3.1				ug/L	1845	Standard
Sr	88	470.0	1.1				ug/L	70	Standard
C	12	286.7	5.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	31.1	36.4				mg/L	12	Standard
Ho-1	165	38.3	7.5				mg/L	12	Standard
Er	166	46.7	44.6				mg/L	20	Standard
I	127	79897.5	13.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		100.068	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		94.936	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606140101
Report Date/Time: Tuesday, June 28, 2016 23:02:15
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.942
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		89.595
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606140101
Report Date/Time: Tuesday, June 28, 2016 23:02:15
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606140201
Sample Date/Time: Tuesday, June 28, 2016 23:03:09
Number of Replicates: 3
Autosampler Position: 351
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76605.0	3.1				ug/L	81760	Standard
	Be	9	21.7	13.3	**0.0135**	0.003	20.6	ug/L	17	Standard
	Al	27	26105502.1	1.1	**180.8194**	3.801	2.1	ug/L	1410	Standard
	Sc	45	39097.0	0.2				ug/L	39543	Standard
	Ti	47	562.0	2.4	**2.2655**	0.057	2.5	ug/L	39	Standard
	V	51	4742.4	3.6	**0.5271**	0.022	4.2	ug/L	1415	Standard
	Cr	52	10082.4	1.2	**0.5007**	0.019	3.8	ug/L	6946	Standard
	Cr	53	12313.3	5.4	**14.7485**	0.907	6.2	ug/L	328	Standard
	Mn	55	1295871.4	1.2	**214.8489**	3.945	1.8	ug/L	823	Standard
	Co	59	1407.4	1.1	**0.1841**	0.002	1.1	ug/L	252	Standard
	Ni	60	4305.3	1.8	**2.3535**	0.034	1.4	ug/L	192	Standard
	Cu	65	1020.0	5.2	**0.4867**	0.034	6.9	ug/L	141	Standard
	Zn	66	18219.9	0.4	**17.0511**	0.173	1.0	ug/L	491	Standard
>	**Ge**	72	478933.9	0.7				ug/L	515087	Standard
	As	75	17040.6	2.0	**16.3399**	0.219	1.3	ug/L	20	Standard
	Se	82	892.6	3.4	**8.2180**	0.229	2.8	ug/L	15	Standard
	Se-1	77	753.4	7.3	**10.0390**	0.842	8.4	ug/L	48	Standard
>	**Ga**	71	218.3	6.6				mg/L	90	Standard
	Rb	85	3558.8	1.6				ug/L	23	Standard
	Y	89	416573.3	1.7				ug/L	448510	Standard
>	**Rh**	103	118.3	21.3				ug/L	15	Standard
	Mo	98	260.0	1.1	**0.0877**	0.004	4.4	ug/L	11	Standard
	Ag	107	80.3	26.9	**-0.0029**	0.004	134.9	ug/L	134	Standard
	Cd	111	40.9	26.1	**0.0184**	0.006	33.2	mg/L	7	Standard
	Cd	114	102.6	27.3	**0.0171**	0.006	37.8	ug/L	16	Standard
>	**In**	115	551143.5	2.5				ug/L	611117	Standard
	Sn	118	835.0	12.0	**0.0548**	0.020	37.1	ug/L	518	Standard
	Sb	123	829.1	3.5	**0.1820**	0.004	2.4	ug/L	127	Standard
	Ba	135	332622.5	0.7	**178.6804**	3.714	2.1	ug/L	45	Standard
	Ce	140	6094.6	2.7				ug/L	298	Standard
>	**Tb**	159	884304.9	2.4				ug/L	977942	Standard
	Ho	165	73.3	20.8				ug/L	12	Standard
	Tl	203	115.3	18.9	**0.0110**	0.003	28.4	ug/L	31	Standard
	Tl	205	95.0	13.9	**0.0127**	0.002	15.2	ug/L	20	Standard
	Pb	206	48021.0	1.3	**10.1649**	0.107	1.1	ug/L	286	Standard
	Pb	207	40222.3	0.8	**9.4401**	0.093	1.0	ug/L	257	Standard
	Pb	208	168471.0	1.1	**10.2223**	0.060	0.6	ug/L	956	Standard
	U	238	714.0	7.4	**0.1118**	0.009	8.4	ug/L	17	Standard
>	**Bi**	209	483931.7	1.1				ug/L	557080	Standard

Sample ID: L1606140201
Report Date/Time: Tuesday, June 28, 2016 23:05:26
Page 1

Approved: June 29, 2016

Na	23	8.3	34.6	**54.1394**	18.829	34.8	mg/L	0	Standard
Mg	24	106065.1	2.2	**142.7401**	3.063	2.1	mg/L	32	Standard
K	39	148.3	8.5	**0.3655**	0.044	12.0	mg/L	30	Standard
Ca	43	363.3	11.2	**50.4710**	6.294	12.5	mg/L	48	Standard
Fe	54	1618.9	1.4	**1.3448**	0.023	1.7	mg/L	159	Standard
Fe	57	835.0	5.7	**2.2263**	0.157	7.1	mg/L	195	Standard
Sc-1	45	39097.0	0.2				mg/L	39543	Standard
Cl	35	36261.3	1.4				ug/L	27642	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	214333.6	4.1				ug/L	1647	Standard
P	31	21892.5	1.8				ug/L	15231	Standard
S	34	2305.2	3.8				ug/L	1845	Standard
Sr	88	363.3	12.3				ug/L	70	Standard
C	12	390.0	0.0				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	133.0	23.5				mg/L	12	Standard
Ho-1	165	73.3	20.8				mg/L	12	Standard
Er	166	76.7	66.9				mg/L	20	Standard
I	127	398539.7	6.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		93.695	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		92.981	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606140201
Report Date/Time: Tuesday, June 28, 2016 23:05:26
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		90.186
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		86.869
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606140201
Report Date/Time: Tuesday, June 28, 2016 23:05:26
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606140401
Sample Date/Time: Tuesday, June 28, 2016 23:06:21
Number of Replicates: 3
Autosampler Position: 352
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74993.5	0.5				ug/L	81760	Standard
	Be	9	21.7	13.3	**0.0141**	0.003	22.5	ug/L	17	Standard
	Al	27	30278310.9	1.4	**214.1402**	2.917	1.4	ug/L	1410	Standard
	Sc	45	39486.4	2.9				ug/L	39543	Standard
	Ti	47	193.3	4.5	**0.6334**	0.048	7.5	ug/L	39	Standard
	V	51	22966.6	2.4	**3.1851**	0.035	1.1	ug/L	1415	Standard
	Cr	52	11319.6	0.7	**0.6125**	0.031	5.1	ug/L	6946	Standard
	Cr	53	1811.8	4.8	**1.7776**	0.107	6.0	ug/L	328	Standard
	Mn	55	8657.1	1.2	**1.2226**	0.034	2.8	ug/L	823	Standard
	Co	59	565.7	7.8	**0.0513**	0.007	13.0	ug/L	252	Standard
	Ni	60	1945.8	2.7	**0.9557**	0.026	2.7	ug/L	192	Standard
	Cu	65	60353.0	0.9	**32.5191**	0.515	1.6	ug/L	141	Standard
	Zn	66	48175.8	1.3	**44.3004**	1.008	2.3	ug/L	491	Standard
>	**Ge**	72	501772.7	1.4				ug/L	515087	Standard
	As	75	367.9	8.8	**0.3760**	0.025	6.7	ug/L	20	Standard
	Se	82	110.4	13.6	**0.8639**	0.120	13.9	ug/L	15	Standard
	Se-1	77	129.0	8.1	**1.0381**	0.123	11.8	ug/L	48	Standard
>	**Ga**	71	68.3	27.7				mg/L	90	Standard
	Rb	85	1198.4	2.3				ug/L	23	Standard
	Y	89	417957.0	1.4				ug/L	448510	Standard
>	**Rh**	103	78.3	35.2				ug/L	15	Standard
	Mo	98	388.3	6.2	**0.1355**	0.009	6.9	ug/L	11	Standard
	Ag	107	94.3	4.8	**-0.0011**	0.001	61.6	ug/L	134	Standard
	Cd	111	42.7	19.5	**0.0186**	0.004	22.3	mg/L	7	Standard
	Cd	114	106.8	24.4	**0.0172**	0.005	31.6	ug/L	16	Standard
>	**In**	115	568064.8	0.3				ug/L	611117	Standard
	Sn	118	741.7	8.4	**0.0330**	0.012	35.5	ug/L	518	Standard
	Sb	123	163.0	3.3	**0.0243**	0.001	4.6	ug/L	127	Standard
	Ba	135	87619.3	0.7	**45.6318**	0.199	0.4	ug/L	45	Standard
	Ce	140	530.0	5.7				ug/L	298	Standard
>	**Tb**	159	902463.1	0.6				ug/L	977942	Standard
	Ho	165	31.7	24.1				ug/L	12	Standard
	Tl	203	89.3	3.6	**0.0068**	0.000	7.0	ug/L	31	Standard
	Tl	205	80.0	18.8	**0.0096**	0.002	23.2	ug/L	20	Standard
	Pb	206	4637.4	0.5	**0.8697**	0.009	1.0	ug/L	286	Standard
	Pb	207	3892.8	1.5	**0.8106**	0.017	2.1	ug/L	257	Standard
	Pb	208	15697.9	1.6	**0.8448**	0.009	1.1	ug/L	956	Standard
	U	238	6211.6	0.5	**0.9400**	0.010	1.1	ug/L	17	Standard
>	**Bi**	209	512163.0	0.7				ug/L	557080	Standard

Sample ID: L1606140401
Report Date/Time: Tuesday, June 28, 2016 23:08:38
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.8762**	18.829	173.1	mg/L	0	Standard
Mg	24	25888.8	0.8	**34.4779**	1.238	3.6	mg/L	32	Standard
K	39	206.7	17.0	**0.5565**	0.104	18.6	mg/L	30	Standard
Ca	43	238.3	20.2	**30.9565**	7.726	25.0	mg/L	48	Standard
Fe	54	118.7	26.0	**-0.0297**	0.026	88.1	mg/L	159	Standard
Fe	57	376.7	12.0	**0.6523**	0.186	28.5	mg/L	195	Standard
> **Sc-1**	45	39486.4	2.9				mg/L	39543	Standard
Cl	35	38202.8	1.2				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	11808.0	7.7				ug/L	1647	Standard
P	31	21151.4	0.3				ug/L	15231	Standard
S	34	2466.9	2.1				ug/L	1845	Standard
Sr	88	291.7	10.5				ug/L	70	Standard
C	12	196.7	43.2				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	28.2	2.0				mg/L	12	Standard
Ho-1	165	31.7	24.1				mg/L	12	Standard
Er	166	36.7	31.5				mg/L	20	Standard
I	127	34317.5	11.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.724	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.415	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606140401
Report Date/Time: Tuesday, June 28, 2016 23:08:38
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		92.955
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		91.937
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606140401
Report Date/Time: Tuesday, June 28, 2016 23:08:38
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606140801
Sample Date/Time: Tuesday, June 28, 2016 23:09:32
Number of Replicates: 3
Autosampler Position: 353
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78174.6	2.2				ug/L	81760	Standard
	Be	9	35.0	24.7	**0.0270**	0.008	30.6	ug/L	17	Standard
	Al	27	18286645.8	1.2	**124.0948**	2.663	2.1	ug/L	1410	Standard
	Sc	45	40988.7	1.5				ug/L	39543	Standard
	Ti	47	225.3	13.4	**0.7985**	0.121	15.1	ug/L	39	Standard
	V	51	51258.6	3.2	**7.6390**	0.153	2.0	ug/L	1415	Standard
	Cr	52	11486.4	2.6	**0.6992**	0.069	9.9	ug/L	6946	Standard
	Cr	53	8145.5	5.8	**9.5287**	0.474	5.0	ug/L	328	Standard
	Mn	55	98520.0	1.7	**16.0208**	0.120	0.7	ug/L	823	Standard
	Co	59	769.4	2.9	**0.0851**	0.002	2.7	ug/L	252	Standard
	Ni	60	1615.4	2.4	**0.8080**	0.032	4.0	ug/L	192	Standard
	Cu	65	1131.0	3.0	**0.5425**	0.014	2.7	ug/L	141	Standard
	Zn	66	4403.0	3.4	**3.4460**	0.103	3.0	ug/L	491	Standard
>	**Ge**	72	484103.1	1.4				ug/L	515087	Standard
	As	75	1970.8	6.0	**1.9047**	0.087	4.5	ug/L	20	Standard
	Se	82	191.5	8.8	**1.6489**	0.132	8.0	ug/L	15	Standard
	Se-1	77	529.7	10.7	**6.7555**	0.697	10.3	ug/L	48	Standard
>	**Ga**	71	85.0	31.1				mg/L	90	Standard
	Rb	85	2531.9	3.1				ug/L	23	Standard
	Y	89	421592.5	4.0				ug/L	448510	Standard
>	**Rh**	103	66.7	30.3				ug/L	15	Standard
	Mo	98	353.3	5.0	**0.1209**	0.007	5.4	ug/L	11	Standard
	Ag	107	105.0	57.1	**0.0006**	0.010	1741.0	ug/L	134	Standard
	Cd	111	58.7	7.1	**0.0265**	0.001	5.4	mg/L	7	Standard
	Cd	114	139.6	12.3	**0.0238**	0.003	13.4	ug/L	16	Standard
>	**In**	115	570844.6	2.8				ug/L	611117	Standard
	Sn	118	760.0	5.7	**0.0358**	0.009	25.3	ug/L	518	Standard
	Sb	123	136.8	41.6	**0.0182**	0.013	71.7	ug/L	127	Standard
	Ba	135	27343.0	2.9	**14.1534**	0.097	0.7	ug/L	45	Standard
	Ce	140	1776.8	2.1				ug/L	298	Standard
>	**Tb**	159	907261.5	0.9				ug/L	977942	Standard
	Ho	165	38.3	32.8				ug/L	12	Standard
	Tl	203	160.7	51.8	**0.0162**	0.011	66.8	ug/L	31	Standard
	Tl	205	123.3	9.4	**0.0162**	0.002	11.0	ug/L	20	Standard
	Pb	206	469.0	16.1	**0.0318**	0.015	47.2	ug/L	286	Standard
	Pb	207	405.3	12.5	**0.0333**	0.011	33.4	ug/L	257	Standard
	Pb	208	1576.7	3.7	**0.0314**	0.003	10.7	ug/L	956	Standard
	U	238	14465.9	3.7	**2.2164**	0.080	3.6	ug/L	17	Standard
>	**Bi**	209	506765.8	0.4				ug/L	557080	Standard

Sample ID: L1606140801
Report Date/Time: Tuesday, June 28, 2016 23:11:49
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**20.4816**	17.733	86.6	mg/L	0	Standard
Mg	24	76429.0	2.1	**98.0879**	0.872	0.9	mg/L	32	Standard
K	39	103.3	20.1	**0.1958**	0.071	36.2	mg/L	30	Standard
Ca	43	268.3	18.7	**34.0467**	7.923	23.3	mg/L	48	Standard
Fe	54	210.5	5.9	**0.0466**	0.009	18.4	mg/L	159	Standard
Fe	57	346.7	13.1	**0.5032**	0.132	26.3	mg/L	195	Standard
Sc-1	45	40988.7	1.5				mg/L	39543	Standard
Cl	35	41040.2	0.8				ug/L	27642	Standard
Kr	83	2.0	173.2				ug/L	2	Standard
Br	81	20580.7	5.3				ug/L	1647	Standard
P	31	21832.4	1.0				ug/L	15231	Standard
S	34	2461.9	6.1				ug/L	1845	Standard
Sr	88	228.3	13.2				ug/L	70	Standard
C	12	196.7	34.6				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	38.1	77.6				mg/L	12	Standard
Ho-1	165	38.3	32.8				mg/L	12	Standard
Er	166	40.0	90.1				mg/L	20	Standard
I	127	105209.9	5.9				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		95.615	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.985	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606140801
Report Date/Time: Tuesday, June 28, 2016 23:11:49
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		93.410
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		90.968
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606140901
Sample Date/Time: Tuesday, June 28, 2016 23:12:44
Number of Replicates: 3
Autosampler Position: 354
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78595.1	3.1				ug/L	81760	Standard
	Be	9	26.7	57.3	**0.0179**	0.015	85.3	ug/L	17	Standard
	Al	27	20167615.6	1.1	**136.1889**	4.854	3.6	ug/L	1410	Standard
	Sc	45	39818.9	1.9				ug/L	39543	Standard
	Ti	47	311.7	10.4	**1.1846**	0.135	11.4	ug/L	39	Standard
	V	51	55787.4	0.6	**8.4391**	0.018	0.2	ug/L	1415	Standard
	Cr	52	16179.6	1.9	**1.4466**	0.047	3.2	ug/L	6946	Standard
	Cr	53	12833.8	2.8	**15.4070**	0.434	2.8	ug/L	328	Standard
	Mn	55	13187.4	1.1	**2.0427**	0.028	1.4	ug/L	823	Standard
	Co	59	670.0	4.7	**0.0714**	0.005	7.2	ug/L	252	Standard
	Ni	60	2540.2	2.3	**1.3479**	0.032	2.4	ug/L	192	Standard
	Cu	65	810.4	1.3	**0.3686**	0.005	1.5	ug/L	141	Standard
	Zn	66	2808.3	3.3	**1.9320**	0.083	4.3	ug/L	491	Standard
>	**Ge**	72	478186.0	0.4				ug/L	515087	Standard
	As	75	1572.1	0.9	**1.5465**	0.010	0.6	ug/L	20	Standard
	Se	82	338.5	3.3	**3.0475**	0.114	3.7	ug/L	15	Standard
	Se-1	77	792.0	6.7	**10.6055**	0.757	7.1	ug/L	48	Standard
>	**Ga**	71	71.7	10.7				mg/L	90	Standard
	Rb	85	3450.4	2.1				ug/L	23	Standard
	Y	89	421286.0	4.0				ug/L	448510	Standard
>	**Rh**	103	76.7	10.0				ug/L	15	Standard
	Mo	98	1179.3	2.2	**0.4594**	0.016	3.6	ug/L	11	Standard
	Ag	107	70.0	6.5	**-0.0048**	0.001	13.6	ug/L	134	Standard
	Cd	111	26.3	13.5	**0.0106**	0.002	18.0	mg/L	7	Standard
	Cd	114	71.5	14.0	**0.0102**	0.002	20.9	ug/L	16	Standard
>	**In**	115	557486.7	1.6				ug/L	611117	Standard
	Sn	118	753.4	2.1	**0.0377**	0.003	7.4	ug/L	518	Standard
	Sb	123	140.1	13.9	**0.0197**	0.005	23.1	ug/L	127	Standard
	Ba	135	30044.1	2.0	**15.9271**	0.116	0.7	ug/L	45	Standard
	Ce	140	1501.7	8.9				ug/L	298	Standard
>	**Tb**	159	887600.8	1.1				ug/L	977942	Standard
	Ho	165	38.3	27.2				ug/L	12	Standard
	Tl	203	136.3	15.5	**0.0135**	0.003	21.6	ug/L	31	Standard
	Tl	205	116.7	8.9	**0.0158**	0.002	10.2	ug/L	20	Standard
	Pb	206	477.7	6.1	**0.0364**	0.006	16.6	ug/L	286	Standard
	Pb	207	398.0	3.8	**0.0344**	0.003	8.7	ug/L	257	Standard
	Pb	208	1707.0	2.5	**0.0420**	0.002	4.6	ug/L	956	Standard
	U	238	23354.0	0.6	**3.6864**	0.026	0.7	ug/L	17	Standard
>	**Bi**	209	492176.6	0.6				ug/L	557080	Standard

Sample ID: L1606140901
Report Date/Time: Tuesday, June 28, 2016 23:15:01
Page 1

Approved: June 29, 2016

Na	23	8.3	124.9	**53.9308**	68.028	126.1	mg/L	0	Standard
Mg	24	78221.4	0.2	**103.3688**	1.796	1.7	mg/L	32	Standard
K	39	181.7	9.7	**0.4684**	0.061	12.9	mg/L	30	Standard
Ca	43	295.0	5.9	**39.1939**	3.495	8.9	mg/L	48	Standard
Fe	54	173.9	5.6	**0.0192**	0.007	35.0	mg/L	159	Standard
Fe	57	403.3	17.1	**0.7302**	0.251	34.4	mg/L	195	Standard
Sc-1	45	39818.9	1.9				mg/L	39543	Standard
Cl	35	38063.1	1.8				ug/L	27642	Standard
Kr	83	1.3	173.2				ug/L	2	Standard
Br	81	39401.2	2.9				ug/L	1647	Standard
P	31	23753.6	2.5				ug/L	15231	Standard
S	34	2233.5	3.9				ug/L	1845	Standard
Sr	88	295.0	6.1				ug/L	70	Standard
C	12	220.0	31.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	222.1	127.8				mg/L	12	Standard
Ho-1	165	38.3	27.2				mg/L	12	Standard
Er	166	26.7	78.1				mg/L	20	Standard
I	127	56189.0	1.6				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		96.129	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		92.836	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606140901
Report Date/Time: Tuesday, June 28, 2016 23:15:01
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		91.224
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		88.349
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606140901
Report Date/Time: Tuesday, June 28, 2016 23:15:01
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 23:15:56
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	73471.1	1.2				ug/L	81760	Standard
	Be	9	47618.1	3.0	**53.3659**	2.229	4.2	ug/L	17	Standard
	Al	27	7287296.4	1.0	**52.5966**	0.249	0.5	ug/L	1410	Standard
	Sc	45	40183.2	3.0				ug/L	39543	Standard
	Ti	47	24656.7	0.7	**99.9368**	1.445	1.4	ug/L	39	Standard
	V	51	342959.6	1.5	**49.6067**	0.741	1.5	ug/L	1415	Standard
	Cr	52	350575.7	1.6	**49.7880**	0.923	1.9	ug/L	6946	Standard
	Cr	53	44859.7	1.2	**51.2482**	0.358	0.7	ug/L	328	Standard
	Mn	55	313688.7	1.9	**48.6980**	0.765	1.6	ug/L	823	Standard
	Co	59	336351.1	1.2	**48.2963**	0.507	1.1	ug/L	252	Standard
	Ni	60	92893.3	2.1	**49.6893**	0.993	2.0	ug/L	192	Standard
	Cu	65	92858.0	0.8	**49.2454**	0.445	0.9	ug/L	141	Standard
	Zn	66	54854.1	1.5	**49.6985**	0.961	1.9	ug/L	491	Standard
>	Ge	72	510232.1	0.7				ug/L	515087	Standard
	As	75	53972.2	1.6	**48.5043**	0.935	1.9	ug/L	20	Standard
	Se	82	5467.9	0.8	**47.8296**	0.724	1.5	ug/L	15	Standard
	Se-1	77	3829.5	1.2	**50.5966**	0.465	0.9	ug/L	48	Standard
>	Ga	71	55.0	32.8				mg/L	90	Standard
	Rb	85	740.0	22.2				ug/L	23	Standard
	Y	89	417471.8	0.9				ug/L	448510	Standard
>	Rh	103	28.3	40.8				ug/L	15	Standard
	Mo	98	255774.9	1.1	**100.4324**	1.068	1.1	ug/L	11	Standard
	Ag	107	314263.4	0.6	**49.6826**	0.310	0.6	ug/L	134	Standard
	Cd	111	99244.0	2.0	**49.2800**	0.642	1.3	mg/L	7	Standard
	Cd	114	244609.6	0.7	**49.9721**	0.117	0.2	ug/L	16	Standard
>	In	115	575695.1	0.7				ug/L	611117	Standard
	Sn	118	277518.7	1.2	**49.7196**	0.393	0.8	ug/L	518	Standard
	Sb	123	223405.3	1.0	**50.2498**	0.315	0.6	ug/L	127	Standard
	Ba	135	95362.2	0.9	**49.0082**	0.194	0.4	ug/L	45	Standard
	Ce	140	66.7	18.9				ug/L	298	Standard
>	Tb	159	900469.8	1.2				ug/L	977942	Standard
	Ho	165	0.0					ug/L	12	Standard
	Tl	203	399704.8	0.8	**50.1537**	0.324	0.6	ug/L	31	Standard
	Tl	205	354566.9	2.1	**51.2307**	0.689	1.3	ug/L	20	Standard
	Pb	206	258134.8	1.1	**50.4360**	0.346	0.7	ug/L	286	Standard
	Pb	207	234508.0	1.0	**50.8014**	0.140	0.3	ug/L	257	Standard
	Pb	208	934488.0	0.8	**52.3272**	0.390	0.7	ug/L	956	Standard
	U	238	348175.4	0.3	**51.3787**	0.514	1.0	ug/L	17	Standard
>	Bi	209	526902.0	1.3				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:18:13
Page 1

Approved: June 29, 2016

Na	23	5.0	100.0	**32.0442**	31.868	99.5	mg/L	0	Standard
Mg	24	3947.2	3.3	**5.1157**	0.206	4.0	mg/L	32	Standard
K	39	1291.7	6.0	**4.1569**	0.313	7.5	mg/L	30	Standard
Ca	43	78.3	13.3	**6.3383**	1.191	18.8	mg/L	48	Standard
Fe	54	5567.6	3.6	**4.8270**	0.308	6.4	mg/L	159	Standard
Fe	57	1653.4	2.1	**4.8734**	0.265	5.4	mg/L	195	Standard
Sc-1	45	40183.2	3.0				mg/L	39543	Standard
Cl	35	36099.6	2.2				ug/L	27642	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	2486.9	19.4				ug/L	1647	Standard
P	31	18199.3	2.9				ug/L	15231	Standard
S	34	2558.5	2.4				ug/L	1845	Standard
Sr	88	88.3	13.1				ug/L	70	Standard
C	12	123.3	18.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	0	Standard
Dy	164	2.5	205.9				mg/L	12	Standard
Ho-1	165	0.0					mg/L	12	Standard
Er	166	16.7	69.3				mg/L	20	Standard
I	127	3410.4	24.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	106.732		
Al	27	105.193		
Sc	45			
Ti	47	99.937		
V	51	99.213		
Cr	52	99.576		
Cr	53			
Mn	55	97.396		
Co	59	96.593		
Ni	60	99.379		
Cu	65	98.491		
Zn	66	99.397		
Ge	72		99.057	
As	75	97.009		
Se	82	95.659		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:18:13
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	100.432		
Ag	107	99.365		
Cd	111	98.560		
Cd	114			
> In	115			94.204
Sn	118	99.439		
Sb	123	100.500		
Ba	135	98.016		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	100.307		
Tl	205			
Pb	206			
Pb	207			
Pb	208	104.654		
U	238	102.757		
> Bi	209			94.583
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:18:13
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, June 28, 2016 23:19:08
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74688.6	0.5				ug/L	81760	Standard
	Be	9	23.3	68.9	**0.0159**	0.018	110.7	ug/L	17	Standard
	Al	27	6963.5	51.4	**0.0350**	0.025	72.0	ug/L	1410	Standard
	Sc	45	39526.5	2.8				ug/L	39543	Standard
	Ti	47	29.3	2.0	**-0.0430**	0.003	6.9	ug/L	39	Standard
	V	51	1022.6	3.0	**-0.0543**	0.006	11.1	ug/L	1415	Standard
	Cr	52	6275.6	1.6	**-0.1243**	0.005	4.1	ug/L	6946	Standard
	Cr	53	655.0	2.0	**0.4308**	0.021	5.0	ug/L	328	Standard
	Mn	55	708.0	10.7	**-0.0355**	0.011	31.8	ug/L	823	Standard
	Co	59	254.0	9.4	**0.0060**	0.003	53.0	ug/L	252	Standard
	Ni	60	190.7	1.6	**-0.0002**	0.001	475.8	ug/L	192	Standard
	Cu	65	144.3	16.1	**-0.0123**	0.012	100.6	ug/L	141	Standard
	Zn	66	441.0	4.3	**-0.4114**	0.015	3.7	ug/L	491	Standard
>	**Ge**	72	497792.7	1.0				ug/L	515087	Standard
	As	75	-15.2	258.6	**0.0262**	0.036	138.8	ug/L	20	Standard
	Se	82	23.5	16.0	**0.0916**	0.033	35.7	ug/L	15	Standard
	Se-1	77	63.7	3.3	**0.1558**	0.037	23.6	ug/L	48	Standard
>	**Ga**	71	50.0	34.6				mg/L	90	Standard
	Rb	85	33.3	56.8				ug/L	23	Standard
	Y	89	425942.8	1.0				ug/L	448510	Standard
>	**Rh**	103	6.7	43.3				ug/L	15	Standard
	Mo	98	281.0	37.4	**0.0876**	0.040	45.9	ug/L	11	Standard
	Ag	107	137.7	21.7	**0.0048**	0.005	95.8	ug/L	134	Standard
	Cd	111	15.2	29.8	**0.0044**	0.002	50.1	mg/L	7	Standard
	Cd	114	43.1	89.6	**0.0036**	0.008	212.7	ug/L	16	Standard
>	**In**	115	595727.8	0.7				ug/L	611117	Standard
	Sn	118	753.4	17.4	**0.0287**	0.023	78.8	ug/L	518	Standard
	Sb	123	441.8	17.1	**0.0832**	0.016	19.8	ug/L	127	Standard
	Ba	135	56.0	54.4	**0.0027**	0.015	556.5	ug/L	45	Standard
	Ce	140	23.3	12.4				ug/L	298	Standard
>	**Tb**	159	931120.0	0.7				ug/L	977942	Standard
	Ho	165	10.0	86.6				ug/L	12	Standard
	Tl	203	88.3	51.6	**0.0060**	0.006	92.0	ug/L	31	Standard
	Tl	205	63.3	63.8	**0.0066**	0.006	85.2	ug/L	20	Standard
	Pb	206	308.0	15.6	**-0.0049**	0.009	177.6	ug/L	286	Standard
	Pb	207	259.7	11.6	**-0.0032**	0.006	187.4	ug/L	257	Standard
	Pb	208	1046.3	7.8	**-0.0033**	0.004	126.3	ug/L	956	Standard
	U	238	64.3	43.1	**0.0063**	0.004	62.8	ug/L	17	Standard
>	**Bi**	209	540861.2	0.7				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 23:21:25
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.5148**	18.203	173.1	mg/L	0	Standard
Mg	24	55.0	41.7	**0.0162**	0.030	187.5	mg/L	32	Standard
K	39	25.0	20.0	**-0.0566**	0.019	34.0	mg/L	30	Standard
Ca	43	33.3	34.6	**-0.2594**	1.827	704.2	mg/L	48	Standard
Fe	54	134.2	10.1	**-0.0157**	0.009	57.1	mg/L	159	Standard
Fe	57	241.7	9.3	**0.1932**	0.098	50.6	mg/L	195	Standard
Sc-1	45	39526.5	2.8				mg/L	39543	Standard
Cl	35	35813.0	1.8				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1830.1	10.1				ug/L	1647	Standard
P	31	18830.0	1.9				ug/L	15231	Standard
S	34	2625.2	3.0				ug/L	1845	Standard
Sr	88	65.0	33.5				ug/L	70	Standard
C	12	80.0	21.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	19.2	93.8				mg/L	12	Standard
Ho-1	165	10.0	86.6				mg/L	12	Standard
Er	166	16.7	91.7				mg/L	20	Standard
I	127	2391.9	12.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.642	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, June 28, 2016 23:21:25
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.482
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.089
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Tuesday, June 28, 2016 23:22:21
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70209.0	1.7				ug/L	81760	Standard
	Be	9	180.0	16.9	**0.2017**	0.039	19.3	ug/L	17	Standard
	Al	27	5133.2	109.4	**0.0240**	0.041	173.1	ug/L	1410	Standard
	Sc	45	36931.6	2.6				ug/L	39543	Standard
	Ti	47	20.7	36.3	**-0.0758**	0.033	44.0	ug/L	39	Standard
	V	51	3515.4	3.3	**0.3352**	0.014	4.2	ug/L	1415	Standard
	Cr	52	11116.7	0.6	**0.6541**	0.028	4.2	ug/L	6946	Standard
	Cr	53	1220.0	4.0	**1.1474**	0.049	4.3	ug/L	328	Standard
	Mn	55	3585.8	2.9	**0.4442**	0.023	5.1	ug/L	823	Standard
	Co	59	2687.2	1.7	**0.3785**	0.013	3.3	ug/L	252	Standard
	Ni	60	2873.6	1.1	**1.5299**	0.031	2.0	ug/L	192	Standard
	Cu	65	1483.4	2.9	**0.7455**	0.015	2.0	ug/L	141	Standard
	Zn	66	6969.9	1.2	**5.9869**	0.185	3.1	ug/L	491	Standard
>	**Ge**	72	480794.8	1.6				ug/L	515087	Standard
	As	75	396.5	4.3	**0.4183**	0.021	5.0	ug/L	20	Standard
	Se	82	61.9	5.1	**0.4561**	0.020	4.4	ug/L	15	Standard
	Se-1	77	78.3	14.0	**0.3969**	0.172	43.4	ug/L	48	Standard
>	**Ga**	71	46.7	32.7				mg/L	90	Standard
	Rb	85	21.7	13.3				ug/L	23	Standard
	Y	89	400775.4	1.5				ug/L	448510	Standard
>	**Rh**	103	13.3	21.7				ug/L	15	Standard
	Mo	98	70.5	1.1	**0.0095**	0.000	1.9	ug/L	11	Standard
	Ag	107	2377.2	2.1	**0.3719**	0.012	3.3	ug/L	134	Standard
	Cd	111	440.9	3.8	**0.2230**	0.005	2.1	mg/L	7	Standard
	Cd	114	1065.8	11.0	**0.2200**	0.026	11.9	ug/L	16	Standard
>	**In**	115	557761.4	1.6				ug/L	611117	Standard
	Sn	118	468.3	18.8	**-0.0150**	0.017	116.5	ug/L	518	Standard
	Sb	123	1709.9	1.5	**0.3843**	0.009	2.4	ug/L	127	Standard
	Ba	135	1358.4	3.9	**0.6957**	0.022	3.2	ug/L	45	Standard
	Ce	140	15.0	33.3				ug/L	298	Standard
>	**Tb**	159	873882.2	1.3				ug/L	977942	Standard
	Ho	165	8.3	124.9				ug/L	12	Standard
	Tl	203	633.3	2.0	**0.0774**	0.002	2.5	ug/L	31	Standard
	Tl	205	590.0	17.7	**0.0858**	0.015	17.2	ug/L	20	Standard
	Pb	206	1230.0	0.6	**0.1853**	0.002	0.8	ug/L	286	Standard
	Pb	207	1060.4	4.9	**0.1798**	0.012	6.4	ug/L	257	Standard
	Pb	208	4496.6	3.9	**0.2002**	0.009	4.3	ug/L	956	Standard
	U	238	2637.6	2.9	**0.3996**	0.008	2.0	ug/L	17	Standard
>	**Bi**	209	509391.4	0.9				ug/L	557080	Standard

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 23:24:38
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	38.3	30.1	**-0.0027**	0.016	586.9	mg/L	32	Standard
K	39	28.3	44.4	**-0.0396**	0.043	109.0	mg/L	30	Standard
Ca	43	26.7	43.3	**-1.0379**	1.747	168.3	mg/L	48	Standard
Fe	54	112.3	22.9	**-0.0284**	0.024	85.0	mg/L	159	Standard
Fe	57	253.3	17.2	**0.2902**	0.150	51.8	mg/L	195	Standard
> **Sc-1**	45	36931.6	2.6				mg/L	39543	Standard
Cl	35	34628.2	2.2				ug/L	27642	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	1663.4	2.8				ug/L	1647	Standard
P	31	13167.4	6.3				ug/L	15231	Standard
S	34	2550.2	3.7				ug/L	1845	Standard
Sr	88	81.7	12.7				ug/L	70	Standard
C	12	96.7	15.8				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	22.7	104.3				mg/L	12	Standard
Ho-1	165	8.3	124.9				mg/L	12	Standard
Er	166	13.3	173.2				mg/L	20	Standard
I	127	1815.1	19.4				mg/L	1162	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	100.854		
	Al	27			
	Sc	45			
	Ti	47			
	V	51	83.798		
	Cr	52	81.760		
	Cr	53			
	Mn	55	88.848		
	Co	59	94.629		
	Ni	60	95.622		
	Cu	65	93.183		
	Zn	66	95.791		
>	Ge	72		93.342	
	As	75	104.574		
	Se	82	114.027		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 23:24:38
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	92.977		
Cd	111	92.929		
Cd	114			
> In	115			91.269
Sn	118			
Sb	123	96.075		
Ba	135	92.766		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	96.789		
Tl	205			
Pb	206			
Pb	207			
Pb	208	100.122		
U	238	99.892		
> Bi	209			91.439
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 8
Report Date/Time: Tuesday, June 28, 2016 23:24:38
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: PBS M4 WG574306-02
Sample Date/Time: Tuesday, June 28, 2016 23:25:34
Number of Replicates: 3
Autosampler Position: 401
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73219.9	2.0				ug/L	81760	Standard
	Be	9	21.7	13.3	**0.0146**	0.003	19.0	ug/L	17	Standard
	Al	27	8870.9	1.6	**0.0499**	0.001	2.1	ug/L	1410	Standard
	Sc	45	39242.4	3.0				ug/L	39543	Standard
	Ti	47	26.0	21.4	**-0.0565**	0.021	37.8	ug/L	39	Standard
	V	51	1183.1	4.4	**-0.0293**	0.008	27.7	ug/L	1415	Standard
	Cr	52	7819.4	0.5	**0.1132**	0.035	31.1	ug/L	6946	Standard
	Cr	53	706.7	5.7	**0.4968**	0.043	8.6	ug/L	328	Standard
	Mn	55	910.4	1.2	**-0.0022**	0.005	247.3	ug/L	823	Standard
	Co	59	210.7	7.4	**-0.0001**	0.003	2281.1	ug/L	252	Standard
	Ni	60	270.7	10.1	**0.0449**	0.018	40.4	ug/L	192	Standard
	Cu	65	197.3	7.2	**0.0174**	0.009	50.7	ug/L	141	Standard
	Zn	66	1637.4	3.9	**0.7282**	0.056	7.7	ug/L	491	Standard
>	**Ge**	72	494675.2	2.7				ug/L	515087	Standard
	As	75	-22.9	72.5	**0.0188**	0.016	85.6	ug/L	20	Standard
	Se	82	19.6	24.9	**0.0572**	0.041	70.8	ug/L	15	Standard
	Se-1	77	57.0	14.0	**0.0674**	0.090	134.0	ug/L	48	Standard
>	**Ga**	71	43.3	17.6				mg/L	90	Standard
	Rb	85	43.3	35.3				ug/L	23	Standard
	Y	89	424600.8	0.9				ug/L	448510	Standard
>	**Rh**	103	15.0	66.7				ug/L	15	Standard
	Mo	98	68.5	24.2	**0.0070**	0.006	90.8	ug/L	11	Standard
	Ag	107	79.3	10.7	**-0.0041**	0.001	32.3	ug/L	134	Standard
	Cd	111	8.5	44.3	**0.0012**	0.002	151.3	mg/L	7	Standard
	Cd	114	29.3	97.6	**0.0009**	0.006	647.6	ug/L	16	Standard
>	**In**	115	594639.0	0.2				ug/L	611117	Standard
	Sn	118	648.3	4.2	**0.0107**	0.005	42.7	ug/L	518	Standard
	Sb	123	146.0	12.4	**0.0190**	0.004	20.5	ug/L	127	Standard
	Ba	135	94.3	2.4	**0.0219**	0.001	5.5	ug/L	45	Standard
	Ce	140	41.7	27.7				ug/L	298	Standard
>	**Tb**	159	926166.4	1.8				ug/L	977942	Standard
	Ho	165	11.7	24.7				ug/L	12	Standard
	Tl	203	20.7	43.9	**-0.0022**	0.001	49.4	ug/L	31	Standard
	Tl	205	15.0	33.3	**-0.0001**	0.001	484.0	ug/L	20	Standard
	Pb	206	313.0	7.9	**-0.0032**	0.004	138.6	ug/L	286	Standard
	Pb	207	286.3	1.1	**0.0033**	0.001	36.5	ug/L	257	Standard
	Pb	208	1144.7	1.9	**0.0029**	0.001	51.3	ug/L	956	Standard
	U	238	10.7	32.9	**-0.0014**	0.001	35.7	ug/L	17	Standard
>	**Bi**	209	533593.4	0.9				ug/L	557080	Standard

Sample ID: PBS M4 WG574306-02
Report Date/Time: Tuesday, June 28, 2016 23:27:51
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	31.7	39.7	**-0.0149**	0.016	104.8	mg/L	32	Standard
K	39	41.7	69.3	**0.0013**	0.101	7981.1	mg/L	30	Standard
Ca	43	18.3	15.7	**-2.5418**	0.422	16.6	mg/L	48	Standard
Fe	54	144.2	9.3	**-0.0052**	0.016	301.7	mg/L	159	Standard
Fe	57	226.7	13.5	**0.1451**	0.083	57.3	mg/L	195	Standard
Sc-1	45	39242.4	3.0				mg/L	39543	Standard
Cl	35	36295.4	2.3				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	2233.5	3.2				ug/L	1647	Standard
P	31	18441.2	3.0				ug/L	15231	Standard
S	34	2560.2	3.7				ug/L	1845	Standard
Sr	88	61.7	26.1				ug/L	70	Standard
C	12	130.0	20.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	22.9	112.2				mg/L	12	Standard
Ho-1	165	11.7	24.7				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	1601.8	24.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		89.555	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.037	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBS M4 WG574306-02
Report Date/Time: Tuesday, June 28, 2016 23:27:51
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.304
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.784
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: LCSS M4 WG574306-03
Sample Date/Time: Tuesday, June 28, 2016 23:28:45
Number of Replicates: 3
Autosampler Position: 402
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72376.0	3.8				ug/L	81760	Standard
	Be	9	23423.1	1.4	**26.6532**	0.807	3.0	ug/L	17	Standard
	Al	27	9074.4	3.8	**0.0521**	0.003	4.9	ug/L	1410	Standard
	Sc	45	39197.3	2.4				ug/L	39543	Standard
	Ti	47	29.3	7.1	**-0.0442**	0.007	16.8	ug/L	39	Standard
	V	51	169885.7	0.8	**24.8594**	0.272	1.1	ug/L	1415	Standard
	Cr	52	181506.5	1.2	**25.6821**	0.081	0.3	ug/L	6946	Standard
	Cr	53	22394.9	2.0	**25.8190**	0.261	1.0	ug/L	328	Standard
	Mn	55	162938.5	1.1	**25.6253**	0.137	0.5	ug/L	823	Standard
	Co	59	171171.6	1.4	**24.9527**	0.363	1.5	ug/L	252	Standard
	Ni	60	47214.1	1.2	**25.6031**	0.067	0.3	ug/L	192	Standard
	Cu	65	47011.2	0.9	**25.2811**	0.042	0.2	ug/L	141	Standard
	Zn	66	28788.3	1.3	**26.1066**	0.165	0.6	ug/L	491	Standard
>	**Ge**	72	502290.9	1.0				ug/L	515087	Standard
	As	75	27387.3	0.6	**25.0208**	0.130	0.5	ug/L	20	Standard
	Se	82	2754.4	0.7	**24.4154**	0.147	0.6	ug/L	15	Standard
	Se-1	77	1916.1	0.9	**25.3637**	0.129	0.5	ug/L	48	Standard
>	**Ga**	71	55.0	18.2				mg/L	90	Standard
	Rb	85	31.7	18.2				ug/L	23	Standard
	Y	89	439399.0	0.9				ug/L	448510	Standard
>	**Rh**	103	31.7	24.1				ug/L	15	Standard
	Mo	98	98.7	7.7	**0.0186**	0.002	12.0	ug/L	11	Standard
	Ag	107	158446.9	0.8	**24.3265**	0.298	1.2	ug/L	134	Standard
	Cd	111	51639.0	0.8	**24.9106**	0.321	1.3	mg/L	7	Standard
	Cd	114	123238.5	0.9	**24.4562**	0.275	1.1	ug/L	16	Standard
>	**In**	115	592677.3	1.9				ug/L	611117	Standard
	Sn	118	653.3	6.2	**0.0119**	0.005	41.0	ug/L	518	Standard
	Sb	123	110361.6	0.7	**24.1094**	0.310	1.3	ug/L	127	Standard
	Ba	135	48745.3	0.9	**24.3246**	0.340	1.4	ug/L	45	Standard
	Ce	140	51.7	39.1				ug/L	298	Standard
>	**Tb**	159	905963.3	1.2				ug/L	977942	Standard
	Ho	165	11.7	65.5				ug/L	12	Standard
	Tl	203	202415.8	0.8	**25.4126**	0.129	0.5	ug/L	31	Standard
	Tl	205	180512.3	0.1	**26.1008**	0.236	0.9	ug/L	20	Standard
	Pb	206	132596.7	0.6	**25.8949**	0.268	1.0	ug/L	286	Standard
	Pb	207	115831.1	1.1	**25.0809**	0.349	1.4	ug/L	257	Standard
	Pb	208	485807.7	1.5	**27.1937**	0.548	2.0	ug/L	956	Standard
	U	238	174705.5	0.6	**25.7950**	0.133	0.5	ug/L	17	Standard
>	**Bi**	209	526543.8	0.9				ug/L	557080	Standard

Sample ID: LCSS M4 WG574306-03
Report Date/Time: Tuesday, June 28, 2016 23:31:02
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	28.3	53.9	**-0.0192**	0.020	104.0	mg/L	32	Standard
K	39	18.3	78.7	**-0.0794**	0.048	60.3	mg/L	30	Standard
Ca	43	26.7	21.7	**-1.2726**	0.799	62.8	mg/L	48	Standard
Fe	54	143.6	13.2	**-0.0059**	0.017	289.5	mg/L	159	Standard
Fe	57	241.7	12.5	**0.2002**	0.121	60.5	mg/L	195	Standard
> **Sc-1**	45	39197.3	2.4				mg/L	39543	Standard
Cl	35	35988.7	2.8				ug/L	27642	Standard
Kr	83	3.3	17.3				ug/L	2	Standard
Br	81	1690.1	11.1				ug/L	1647	Standard
P	31	17910.6	1.0				ug/L	15231	Standard
S	34	2548.5	6.5				ug/L	1845	Standard
Sr	88	75.0	24.0				ug/L	70	Standard
C	12	120.0	16.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	16.5	68.3				mg/L	12	Standard
Ho-1	165	11.7	65.5				mg/L	12	Standard
Er	166	3.3	173.2				mg/L	20	Standard
I	127	1618.4	15.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		88.522	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.516	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSS M4 WG574306-03
Report Date/Time: Tuesday, June 28, 2016 23:31:02
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.983
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.518
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSS M4 WG574306-03
Report Date/Time: Tuesday, June 28, 2016 23:31:02
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606129707 WG574306-01
Sample Date/Time: Tuesday, June 28, 2016 23:31:57
Number of Replicates: 3
Autosampler Position: 403
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73039.0	1.3				ug/L	81760	Standard
	Be	9	626.7	10.9	**0.6973**	0.086	12.3	ug/L	17	Standard
	Al	27	6016122.3	2.0	**43.6794**	0.992	2.3	ug/L	1410	Standard
	Sc	45	43592.6	1.8				ug/L	39543	Standard
	Ti	47	1645.1	1.5	**6.4955**	0.085	1.3	ug/L	39	Standard
	V	51	103833.7	2.4	**14.8338**	0.288	1.9	ug/L	1415	Standard
	Cr	52	59851.7	1.4	**7.6005**	0.153	2.0	ug/L	6946	Standard
	Cr	53	7268.4	3.1	**7.9966**	0.305	3.8	ug/L	328	Standard
	Mn	55	3268160.2	1.2	**507.4707**	9.760	1.9	ug/L	823	Standard
	Co	59	39533.2	0.9	**5.6343**	0.098	1.7	ug/L	252	Standard
	Ni	60	18874.4	2.7	**9.9863**	0.299	3.0	ug/L	192	Standard
	Cu	65	69246.2	1.1	**36.6021**	0.305	0.8	ug/L	141	Standard
	Zn	66	86637.7	0.8	**78.7605**	0.745	0.9	ug/L	491	Standard
>	**Ge**	72	511586.4	0.8				ug/L	515087	Standard
	As	75	4823.9	1.5	**4.3600**	0.029	0.7	ug/L	20	Standard
	Se	82	100.3	6.0	**0.7578**	0.047	6.2	ug/L	15	Standard
	Se-1	77	133.3	1.1	**1.0631**	0.006	0.6	ug/L	48	Standard
>	**Ga**	71	7216.7	0.8				mg/L	90	Standard
	Rb	85	77067.2	0.1				ug/L	23	Standard
	Y	89	607493.8	2.8				ug/L	448510	Standard
>	**Rh**	103	25.0	34.6				ug/L	15	Standard
	Mo	98	512.0	2.8	**0.1759**	0.005	2.9	ug/L	11	Standard
	Ag	107	750.4	6.7	**0.0988**	0.008	7.8	ug/L	134	Standard
	Cd	111	815.5	1.0	**0.3898**	0.005	1.3	mg/L	7	Standard
	Cd	114	2038.2	2.1	**0.3989**	0.010	2.4	ug/L	16	Standard
>	**In**	115	593721.9	0.4				ug/L	611117	Standard
	Sn	118	941.7	10.4	**0.0620**	0.016	26.6	ug/L	518	Standard
	Sb	123	1005.8	16.4	**0.2066**	0.037	17.7	ug/L	127	Standard
	Ba	135	393236.3	1.1	**196.0335**	2.433	1.2	ug/L	45	Standard
	Ce	140	857283.3	2.5				ug/L	298	Standard
>	**Tb**	159	944884.3	0.7				ug/L	977942	Standard
	Ho	165	15139.2	2.1				ug/L	12	Standard
	Tl	203	479.0	10.1	**0.0546**	0.006	10.6	ug/L	31	Standard
	Tl	205	396.7	12.4	**0.0544**	0.008	14.1	ug/L	20	Standard
	Pb	206	209606.6	0.8	**40.4302**	0.428	1.1	ug/L	286	Standard
	Pb	207	173972.5	0.8	**37.2011**	0.310	0.8	ug/L	257	Standard
	Pb	208	734775.6	0.6	**40.6204**	0.824	2.0	ug/L	956	Standard
	U	238	3055.0	2.6	**0.4423**	0.014	3.2	ug/L	17	Standard
>	**Bi**	209	533598.4	1.5				ug/L	557080	Standard

Sample ID: L1606129707 WG574306-01
Report Date/Time: Tuesday, June 28, 2016 23:34:13
Page 1

Approved: June 29, 2016

Na	23	3.3	173.2	**19.2926**	33.407	173.2	mg/L	0	Standard
Mg	24	75.0	26.7	**0.0334**	0.024	70.8	mg/L	32	Standard
K	39	210.0	4.8	**0.5021**	0.023	4.7	mg/L	30	Standard
Ca	43	75.0	29.1	**4.9537**	2.848	57.5	mg/L	48	Standard
Fe	54	6552.0	2.8	**5.2441**	0.230	4.4	mg/L	159	Standard
Fe	57	1841.8	2.3	**5.0156**	0.091	1.8	mg/L	195	Standard
> **Sc-1**	45	43592.6	1.8				mg/L	39543	Standard
Cl	35	34528.0	1.5				ug/L	27642	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	3080.3	6.6				ug/L	1647	Standard
P	31	18262.7	4.8				ug/L	15231	Standard
S	34	2550.2	4.9				ug/L	1845	Standard
Sr	88	106.7	7.2				ug/L	70	Standard
C	12	370.0	20.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	40.0				mg/L	0	Standard
Dy	164	25551.2	3.2				mg/L	12	Standard
Ho-1	165	15139.2	2.1				mg/L	12	Standard
Er	166	14156.6	1.1				mg/L	20	Standard
I	127	14104.9	1.1				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		89.333	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.320	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129707 WG574306-01
Report Date/Time: Tuesday, June 28, 2016 23:34:13
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.154
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.785
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606129707 WG574306-01
Report Date/Time: Tuesday, June 28, 2016 23:34:13
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606129708S WG574306-04
Sample Date/Time: Tuesday, June 28, 2016 23:35:08
Number of Replicates: 3
Autosampler Position: 404
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77425.7	0.5				ug/L	81760	Standard
	Be	9	24978.9	2.1	**26.5514**	0.597	2.2	ug/L	17	Standard
	Al	27	20723622.4	0.1	**141.9584**	0.685	0.5	ug/L	1410	Standard
	Sc	45	47225.2	3.2				ug/L	39543	Standard
	Ti	47	4499.3	2.9	**18.2724**	0.558	3.1	ug/L	39	Standard
	V	51	344846.0	0.3	**50.3538**	0.699	1.4	ug/L	1415	Standard
	Cr	52	312957.5	0.8	**44.7646**	1.096	2.4	ug/L	6946	Standard
	Cr	53	39581.6	1.4	**45.6150**	1.229	2.7	ug/L	328	Standard
	Mn	55	5573425.5	1.4	**875.8629**	6.350	0.7	ug/L	823	Standard
	Co	59	254404.2	1.0	**36.8653**	0.333	0.9	ug/L	252	Standard
	Ni	60	98601.5	0.1	**53.2500**	0.840	1.6	ug/L	192	Standard
	Cu	65	83052.3	0.8	**44.4536**	0.827	1.9	ug/L	141	Standard
	Zn	66	91760.7	0.8	**84.4843**	0.972	1.2	ug/L	491	Standard
>	**Ge**	72	505521.9	1.6				ug/L	515087	Standard
	As	75	29384.3	0.4	**26.6737**	0.341	1.3	ug/L	20	Standard
	Se	82	1902.3	1.0	**16.7176**	0.177	1.1	ug/L	15	Standard
	Se-1	77	1368.1	4.0	**17.7813**	0.553	3.1	ug/L	48	Standard
>	**Ga**	71	22081.1	1.8				mg/L	90	Standard
	Rb	85	217862.7	3.0				ug/L	23	Standard
	Y	89	681158.2	0.5				ug/L	448510	Standard
>	**Rh**	103	41.7	13.9				ug/L	15	Standard
	Mo	98	502.0	0.9	**0.1763**	0.003	1.6	ug/L	11	Standard
	Ag	107	151390.6	0.2	**23.7106**	0.326	1.4	ug/L	134	Standard
	Cd	111	50168.4	0.1	**24.6878**	0.285	1.2	mg/L	7	Standard
	Cd	114	122073.8	0.4	**24.7133**	0.372	1.5	ug/L	16	Standard
>	**In**	115	580961.0	1.3				ug/L	611117	Standard
	Sn	118	790.0	7.7	**0.0385**	0.009	23.5	ug/L	518	Standard
	Sb	123	3165.4	5.9	**0.6928**	0.037	5.4	ug/L	127	Standard
	Ba	135	532338.7	0.4	**271.2464**	4.110	1.5	ug/L	45	Standard
	Ce	140	1309534.5	1.1				ug/L	298	Standard
>	**Tb**	159	945803.3	0.3				ug/L	977942	Standard
	Ho	165	21665.5	2.8				ug/L	12	Standard
	Tl	203	195988.5	0.7	**24.3661**	0.181	0.7	ug/L	31	Standard
	Tl	205	172319.6	2.0	**24.6721**	0.452	1.8	ug/L	20	Standard
	Pb	206	254741.0	0.7	**49.3204**	0.423	0.9	ug/L	286	Standard
	Pb	207	216152.1	1.4	**46.3962**	0.722	1.6	ug/L	257	Standard
	Pb	208	888930.7	0.3	**49.3200**	0.099	0.2	ug/L	956	Standard
	U	238	171966.4	0.9	**25.1430**	0.188	0.7	ug/L	17	Standard
>	**Bi**	209	531711.1	0.2				ug/L	557080	Standard

Sample ID: L1606129708S WG574306-04
Report Date/Time: Tuesday, June 28, 2016 23:37:25
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	125.0	34.9	**0.0812**	0.045	55.3	mg/L	32	Standard
K	39	518.3	8.6	**1.3288**	0.177	13.3	mg/L	30	Standard
Ca	43	238.3	9.5	**24.9959**	3.131	12.5	mg/L	48	Standard
Fe	54	18930.5	0.8	**14.2167**	0.407	2.9	mg/L	159	Standard
Fe	57	5154.2	3.0	**13.9467**	0.213	1.5	mg/L	195	Standard
> **Sc-1**	45	47225.2	3.2				mg/L	39543	Standard
Cl	35	33388.1	1.2				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	2326.8	4.2				ug/L	1647	Standard
P	31	19087.0	1.5				ug/L	15231	Standard
S	34	2648.6	1.1				ug/L	1845	Standard
Sr	88	121.7	17.1				ug/L	70	Standard
C	12	206.7	36.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	36.7	56.8				mg/L	0	Standard
Dy	164	35480.3	0.5				mg/L	12	Standard
Ho-1	165	21665.5	2.8				mg/L	12	Standard
Er	166	19556.0	2.5				mg/L	20	Standard
I	127	13694.5	2.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.699	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.143	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129708S WG574306-04
Report Date/Time: Tuesday, June 28, 2016 23:37:25
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.065
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.446
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606129708S WG574306-04
Report Date/Time: Tuesday, June 28, 2016 23:37:25
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606129709SD WG574306-05
Sample Date/Time: Tuesday, June 28, 2016 23:38:19
Number of Replicates: 3
Autosampler Position: 405
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	74161.1	1.1				ug/L	81760	Standard
	Be	9	23705.2	3.8	**26.3039**	0.872	3.3	ug/L	17	Standard
	Al	27	6111125.0	1.3	**43.6951**	0.516	1.2	ug/L	1410	Standard
	Sc	45	44620.6	1.2				ug/L	39543	Standard
	Ti	47	2932.6	3.4	**11.8635**	0.492	4.1	ug/L	39	Standard
	V	51	246865.2	0.2	**36.0158**	0.450	1.2	ug/L	1415	Standard
	Cr	52	213059.8	1.6	**30.1549**	0.173	0.6	ug/L	6946	Standard
	Cr	53	26094.1	1.0	**29.9771**	0.650	2.2	ug/L	328	Standard
	Mn	55	1521366.7	1.6	**239.1613**	2.815	1.2	ug/L	823	Standard
	Co	59	194513.2	1.8	**28.2005**	0.410	1.5	ug/L	252	Standard
	Ni	60	61728.2	1.1	**33.3195**	0.163	0.5	ug/L	192	Standard
	Cu	65	74436.8	1.3	**39.8619**	0.705	1.8	ug/L	141	Standard
	Zn	66	58904.3	0.8	**53.9789**	0.545	1.0	ug/L	491	Standard
>	Ge	72	505111.3	1.2				ug/L	515087	Standard
	As	75	31837.0	1.5	**28.9187**	0.503	1.7	ug/L	20	Standard
	Se	82	2328.1	1.1	**20.5054**	0.470	2.3	ug/L	15	Standard
	Se-1	77	1619.4	1.4	**21.2042**	0.440	2.1	ug/L	48	Standard
>	Ga	71	6334.7	3.0				mg/L	90	Standard
	Rb	85	74980.1	1.6				ug/L	23	Standard
	Y	89	605551.0	0.9				ug/L	448510	Standard
>	Rh	103	31.7	39.7				ug/L	15	Standard
	Mo	98	404.4	5.7	**0.1335**	0.008	5.8	ug/L	11	Standard
	Ag	107	149725.6	1.6	**22.7377**	0.610	2.7	ug/L	134	Standard
	Cd	111	49979.3	1.2	**23.8462**	0.386	1.6	mg/L	7	Standard
	Cd	114	121386.1	1.5	**23.8269**	0.554	2.3	ug/L	16	Standard
>	In	115	599198.2	1.3				ug/L	611117	Standard
	Sn	118	811.7	7.5	**0.0382**	0.012	32.5	ug/L	518	Standard
	Sb	123	19138.7	1.5	**4.1252**	0.117	2.8	ug/L	127	Standard
	Ba	135	258930.8	0.8	**127.9040**	1.896	1.5	ug/L	45	Standard
	Ce	140	760518.7	0.8				ug/L	298	Standard
>	Tb	159	951283.8	0.9				ug/L	977942	Standard
	Ho	165	14106.6	0.1				ug/L	12	Standard
	Tl	203	188730.2	0.5	**23.1187**	0.087	0.4	ug/L	31	Standard
	Tl	205	170134.9	1.0	**24.0027**	0.337	1.4	ug/L	20	Standard
	Pb	206	208981.3	1.7	**39.8515**	0.361	0.9	ug/L	286	Standard
	Pb	207	178505.4	0.4	**37.7421**	0.330	0.9	ug/L	257	Standard
	Pb	208	755416.1	1.2	**41.2868**	0.479	1.2	ug/L	956	Standard
	U	238	170025.1	1.0	**24.4935**	0.095	0.4	ug/L	17	Standard
>	Bi	209	539649.3	0.9				ug/L	557080	Standard

Sample ID: L1606129709SD WG574306-05
Report Date/Time: Tuesday, June 28, 2016 23:40:36
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**19.0051**	16.458	86.6	mg/L	0	Standard
Mg	24	83.3	27.7	**0.0410**	0.026	64.1	mg/L	32	Standard
K	39	181.7	7.9	**0.4027**	0.045	11.2	mg/L	30	Standard
Ca	43	75.0	26.7	**4.7351**	2.622	55.4	mg/L	48	Standard
Fe	54	5650.5	3.5	**4.3963**	0.211	4.8	mg/L	159	Standard
Fe	57	1623.4	7.9	**4.2316**	0.365	8.6	mg/L	195	Standard
> **Sc-1**	45	44620.6	1.2				mg/L	39543	Standard
Cl	35	32136.0	1.3				ug/L	27642	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1900.1	7.1				ug/L	1647	Standard
P	31	17840.5	2.2				ug/L	15231	Standard
S	34	2363.5	5.1				ug/L	1845	Standard
Sr	88	98.3	20.5				ug/L	70	Standard
C	12	150.0	37.1				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	20.0				mg/L	0	Standard
Dy	164	23948.7	0.5				mg/L	12	Standard
Ho-1	165	14106.6	0.1				mg/L	12	Standard
Er	166	12898.8	4.1				mg/L	20	Standard
I	127	5567.7	3.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		90.706	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.063	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129709SD WG574306-05
Report Date/Time: Tuesday, June 28, 2016 23:40:36
Page 2

Approved: June 29, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.050
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	96.871
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606129709SD WG574306-05
Report Date/Time: Tuesday, June 28, 2016 23:40:36
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606144301
Sample Date/Time: Tuesday, June 28, 2016 23:41:31
Number of Replicates: 3
Autosampler Position: 406
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72988.8	2.0				ug/L	81760	Standard
	Be	9	916.7	4.9	**1.0251**	0.070	6.9	ug/L	17	Standard
	Al	27	29327581.2	1.7	**213.1994**	7.066	3.3	ug/L	1410	Standard
	Sc	45	42180.3	1.1				ug/L	39543	Standard
	Ti	47	15217.0	2.1	**62.4530**	1.008	1.6	ug/L	39	Standard
	V	51	58550.8	0.3	**8.4144**	0.093	1.1	ug/L	1415	Standard
	Cr	52	50375.3	2.1	**6.3484**	0.119	1.9	ug/L	6946	Standard
	Cr	53	5892.8	2.4	**6.5282**	0.148	2.3	ug/L	328	Standard
	Mn	55	1501608.8	1.0	**236.8837**	1.387	0.6	ug/L	823	Standard
	Co	59	11082.4	1.1	**1.5829**	0.023	1.4	ug/L	252	Standard
	Ni	60	9938.9	1.0	**5.2962**	0.105	2.0	ug/L	192	Standard
	Cu	65	12880.8	1.6	**6.8461**	0.032	0.5	ug/L	141	Standard
	Zn	66	52551.7	1.1	**48.2375**	0.337	0.7	ug/L	491	Standard
>	**Ge**	72	503346.2	1.1				ug/L	515087	Standard
	As	75	9638.5	1.8	**8.8128**	0.107	1.2	ug/L	20	Standard
	Se	82	89.1	11.5	**0.6728**	0.088	13.1	ug/L	15	Standard
	Se-1	77	112.0	20.6	**0.8047**	0.328	40.8	ug/L	48	Standard
>	**Ga**	71	5402.6	1.6				mg/L	90	Standard
	Rb	85	27092.5	2.9				ug/L	23	Standard
	Y	89	512619.4	0.7				ug/L	448510	Standard
>	**Rh**	103	40.0	45.1				ug/L	15	Standard
	Mo	98	2129.4	2.5	**0.7922**	0.032	4.0	ug/L	11	Standard
	Ag	107	397.3	28.5	**0.0446**	0.016	36.8	ug/L	134	Standard
	Cd	111	847.6	0.8	**0.4054**	0.009	2.1	mg/L	7	Standard
	Cd	114	2047.7	3.4	**0.4008**	0.011	2.7	ug/L	16	Standard
>	**In**	115	593663.8	1.5				ug/L	611117	Standard
	Sn	118	1955.1	5.4	**0.2386**	0.021	8.9	ug/L	518	Standard
	Sb	123	857.9	5.2	**0.1744**	0.012	7.0	ug/L	127	Standard
	Ba	135	91007.2	1.4	**45.3648**	1.263	2.8	ug/L	45	Standard
	Ce	140	427372.5	1.2				ug/L	298	Standard
>	**Tb**	159	914890.2	3.3				ug/L	977942	Standard
	Ho	165	5369.3	2.3				ug/L	12	Standard
	Tl	203	3260.4	2.3	**0.4133**	0.015	3.6	ug/L	31	Standard
	Tl	205	2856.9	4.2	**0.4193**	0.011	2.7	ug/L	20	Standard
	Pb	206	241551.6	0.7	**48.2276**	1.641	3.4	ug/L	286	Standard
	Pb	207	199984.5	0.7	**44.2621**	1.381	3.1	ug/L	257	Standard
	Pb	208	845904.8	0.5	**48.3946**	1.395	2.9	ug/L	956	Standard
	U	238	4244.6	0.7	**0.6370**	0.016	2.5	ug/L	17	Standard
>	**Bi**	209	515907.9	2.8				ug/L	557080	Standard

Sample ID: L1606144301
Report Date/Time: Tuesday, June 28, 2016 23:43:48
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.1607**	17.590	173.1	mg/L	0	Standard
Mg	24	176.7	13.4	**0.1635**	0.031	19.2	mg/L	32	Standard
K	39	111.7	29.8	**0.2126**	0.107	50.1	mg/L	30	Standard
Ca	43	486.7	20.1	**63.9109**	13.520	21.2	mg/L	48	Standard
Fe	54	4570.1	3.9	**3.7396**	0.111	3.0	mg/L	159	Standard
Fe	57	1611.8	3.6	**4.4774**	0.217	4.9	mg/L	195	Standard
Sc-1	45	42180.3	1.1				mg/L	39543	Standard
Cl	35	34952.3	0.5				ug/L	27642	Standard
Kr	83	2.3	65.5				ug/L	2	Standard
Br	81	2820.3	6.2				ug/L	1647	Standard
P	31	18733.3	7.5				ug/L	15231	Standard
S	34	2421.9	7.6				ug/L	1845	Standard
Sr	88	105.0	16.5				ug/L	70	Standard
C	12	266.7	9.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	83.3	60.4				mg/L	0	Standard
Dy	164	8334.8	4.4				mg/L	12	Standard
Ho-1	165	5369.3	2.3				mg/L	12	Standard
Er	166	5100.9	4.8				mg/L	20	Standard
I	127	10480.3	8.4				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		89.272	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.721	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144301
Report Date/Time: Tuesday, June 28, 2016 23:43:48
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.144
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		92.609
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606144302
Sample Date/Time: Tuesday, June 28, 2016 23:44:42
Number of Replicates: 3
Autosampler Position: 407
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	71243.8	1.4				ug/L	81760	Standard
	Be	9	1080.0	8.5	**1.2377**	0.094	7.6	ug/L	17	Standard
	Al	27	5504266.2	2.7	**40.9647**	0.928	2.3	ug/L	1410	Standard
	Sc	45	42188.7	3.4				ug/L	39543	Standard
	Ti	47	15737.8	0.6	**62.0152**	0.652	1.1	ug/L	39	Standard
	V	51	45512.4	1.5	**6.2276**	0.158	2.5	ug/L	1415	Standard
	Cr	52	44821.9	1.8	**5.2702**	0.176	3.3	ug/L	6946	Standard
	Cr	53	5345.9	2.2	**5.6430**	0.187	3.3	ug/L	328	Standard
	Mn	55	1510551.7	1.4	**228.8078**	5.482	2.4	ug/L	823	Standard
	Co	59	8314.6	0.3	**1.1314**	0.011	1.0	ug/L	252	Standard
	Ni	60	7372.1	1.0	**3.7410**	0.046	1.2	ug/L	192	Standard
	Cu	65	13129.7	1.4	**6.6988**	0.146	2.2	ug/L	141	Standard
	Zn	66	50712.1	0.8	**44.6331**	0.723	1.6	ug/L	491	Standard
>	**Ge**	72	524277.6	1.0				ug/L	515087	Standard
	As	75	7848.1	0.4	**6.8989**	0.093	1.3	ug/L	20	Standard
	Se	82	94.2	13.2	**0.6849**	0.109	15.9	ug/L	15	Standard
	Se-1	77	110.0	9.5	**0.7149**	0.123	17.2	ug/L	48	Standard
>	**Ga**	71	4740.7	7.0				mg/L	90	Standard
	Rb	85	22735.4	3.5				ug/L	23	Standard
	Y	89	510942.2	1.8				ug/L	448510	Standard
>	**Rh**	103	18.3	110.2				ug/L	15	Standard
	Mo	98	3616.2	2.9	**1.3869**	0.053	3.8	ug/L	11	Standard
	Ag	107	415.3	27.2	**0.0487**	0.017	35.1	ug/L	134	Standard
	Cd	111	879.6	1.3	**0.4294**	0.003	0.8	mg/L	7	Standard
	Cd	114	2156.4	4.6	**0.4311**	0.017	3.9	ug/L	16	Standard
>	**In**	115	581659.3	1.0				ug/L	611117	Standard
	Sn	118	1941.8	8.1	**0.2429**	0.025	10.2	ug/L	518	Standard
	Sb	123	795.2	3.5	**0.1642**	0.005	3.1	ug/L	127	Standard
	Ba	135	105670.2	0.5	**53.7566**	0.798	1.5	ug/L	45	Standard
	Ce	140	489350.3	0.7				ug/L	298	Standard
>	**Tb**	159	914906.3	0.9				ug/L	977942	Standard
	Ho	165	5852.8	3.2				ug/L	12	Standard
	Tl	203	2193.8	7.0	**0.2695**	0.018	6.5	ug/L	31	Standard
	Tl	205	1858.4	8.2	**0.2655**	0.024	8.9	ug/L	20	Standard
	Pb	206	250441.6	1.0	**48.7728**	1.107	2.3	ug/L	286	Standard
	Pb	207	208820.1	0.2	**45.0828**	0.807	1.8	ug/L	257	Standard
	Pb	208	871187.9	1.2	**48.6248**	1.437	3.0	ug/L	956	Standard
	U	238	4748.1	0.9	**0.6954**	0.016	2.3	ug/L	17	Standard
>	**Bi**	209	528719.9	1.7				ug/L	557080	Standard

Sample ID: L1606144302
Report Date/Time: Tuesday, June 28, 2016 23:46:59
Page 1

Approved: June 29, 2016

Na	23	6.7	43.3	**39.8641**	16.152	40.5	mg/L	0	Standard
Mg	24	568.3	2.7	**0.6523**	0.023	3.6	mg/L	32	Standard
K	39	88.3	11.8	**0.1385**	0.034	24.4	mg/L	30	Standard
Ca	43	236.7	18.2	**28.4335**	6.608	23.2	mg/L	48	Standard
Fe	54	3607.3	1.4	**2.9257**	0.132	4.5	mg/L	159	Standard
Fe	57	1208.4	3.1	**3.2044**	0.245	7.6	mg/L	195	Standard
> **Sc-1**	45	42188.7	3.4				mg/L	39543	Standard
Cl	35	33795.0	1.6				ug/L	27642	Standard
Kr	83	3.3	75.5				ug/L	2	Standard
Br	81	2200.2	4.1				ug/L	1647	Standard
P	31	17892.2	3.9				ug/L	15231	Standard
S	34	2398.5	0.8				ug/L	1845	Standard
Sr	88	86.7	23.3				ug/L	70	Standard
C	12	246.7	15.3				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	76.7	30.1				mg/L	0	Standard
Dy	164	8934.1	5.5				mg/L	12	Standard
Ho-1	165	5852.8	3.2				mg/L	12	Standard
Er	166	5817.8	2.9				mg/L	20	Standard
I	127	10927.3	3.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		87.138	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.784	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144302
Report Date/Time: Tuesday, June 28, 2016 23:46:59
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.180
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.909
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606144302
Report Date/Time: Tuesday, June 28, 2016 23:46:59
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606144302PS WG574573-01
Sample Date/Time: Tuesday, June 28, 2016 23:47:54
Number of Replicates: 3
Autosampler Position: 408
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74517.8	1.0				ug/L	81760	Standard
	Be	9	49025.9	0.9	**54.1590**	0.839	1.5	ug/L	17	Standard
	Al	27	5488864.5	1.0	**39.0606**	0.777	2.0	ug/L	1410	Standard
	Sc	45	43992.1	1.6				ug/L	39543	Standard
	Ti	47	15590.7	1.3	**62.0794**	2.775	4.5	ug/L	39	Standard
	V	51	384779.8	0.6	**54.7401**	1.764	3.2	ug/L	1415	Standard
	Cr	52	400124.0	1.3	**56.0091**	2.445	4.4	ug/L	6946	Standard
	Cr	53	50664.6	1.6	**56.9262**	0.956	1.7	ug/L	328	Standard
	Mn	55	1785071.4	0.4	**273.1709**	9.729	3.6	ug/L	823	Standard
	Co	59	343226.4	0.5	**48.4551**	1.560	3.2	ug/L	252	Standard
	Ni	60	101539.8	1.7	**53.4214**	2.390	4.5	ug/L	192	Standard
	Cu	65	107630.3	0.5	**56.1285**	1.666	3.0	ug/L	141	Standard
	Zn	66	105446.7	0.4	**94.6554**	2.823	3.0	ug/L	491	Standard
>	**Ge**	72	519316.9	3.3				ug/L	515087	Standard
	As	75	63635.8	1.4	**56.2203**	2.024	3.6	ug/L	20	Standard
	Se	82	5672.0	2.3	**48.8026**	2.645	5.4	ug/L	15	Standard
	Se-1	77	3855.8	0.1	**50.0819**	1.642	3.3	ug/L	48	Standard
>	**Ga**	71	4749.1	6.7				mg/L	90	Standard
	Rb	85	22398.2	1.3				ug/L	23	Standard
	Y	89	511504.7	2.1				ug/L	448510	Standard
>	**Rh**	103	58.3	34.6				ug/L	15	Standard
	Mo	98	3550.9	0.5	**1.3279**	0.015	1.1	ug/L	11	Standard
	Ag	107	317741.2	0.8	**48.5166**	0.462	1.0	ug/L	134	Standard
	Cd	111	103027.6	1.2	**49.4156**	0.686	1.4	mg/L	7	Standard
	Cd	114	250223.1	1.7	**49.3820**	1.399	2.8	ug/L	16	Standard
>	**In**	115	596094.6	1.6				ug/L	611117	Standard
	Sn	118	2016.8	5.9	**0.2476**	0.017	6.7	ug/L	518	Standard
	Sb	123	226110.3	0.9	**49.1269**	0.970	2.0	ug/L	127	Standard
	Ba	135	199376.6	0.4	**98.9961**	1.356	1.4	ug/L	45	Standard
	Ce	140	473631.7	1.6				ug/L	298	Standard
>	**Tb**	159	946510.9	0.6				ug/L	977942	Standard
	Ho	165	5706.1	1.8				ug/L	12	Standard
	Tl	203	404565.2	1.6	**49.2192**	0.432	0.9	ug/L	31	Standard
	Tl	205	358549.4	2.4	**50.2279**	0.323	0.6	ug/L	20	Standard
	Pb	206	506766.3	0.6	**96.0764**	1.688	1.8	ug/L	286	Standard
	Pb	207	430425.5	0.4	**90.4686**	1.648	1.8	ug/L	257	Standard
	Pb	208	1789019.3	1.0	**97.1851**	0.775	0.8	ug/L	956	Standard
	U	238	348425.0	0.3	**49.8561**	0.765	1.5	ug/L	17	Standard
>	**Bi**	209	543434.9	1.8				ug/L	557080	Standard

Sample ID: L1606144302PS WG574573-01
Report Date/Time: Tuesday, June 28, 2016 23:50:11
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**19.2423**	16.666	86.6	mg/L	0	Standard
Mg	24	530.0	3.8	**0.5773**	0.030	5.2	mg/L	32	Standard
K	39	95.0	5.3	**0.1471**	0.014	9.3	mg/L	30	Standard
Ca	43	235.0	26.6	**26.8058**	8.969	33.5	mg/L	48	Standard
Fe	54	3632.6	2.6	**2.8189**	0.112	4.0	mg/L	159	Standard
Fe	57	1290.1	7.8	**3.2920**	0.337	10.2	mg/L	195	Standard
> **Sc-1**	45	43992.1	1.6				mg/L	39543	Standard
Cl	35	33127.5	1.3				ug/L	27642	Standard
Kr	83	2.0	100.0				ug/L	2	Standard
Br	81	2050.1	2.6				ug/L	1647	Standard
P	31	18372.8	5.1				ug/L	15231	Standard
S	34	2410.2	0.9				ug/L	1845	Standard
Sr	88	108.3	17.5				ug/L	70	Standard
C	12	236.7	47.5				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	86.7	70.5				mg/L	0	Standard
Dy	164	8282.4	2.8				mg/L	12	Standard
Ho-1	165	5706.1	1.8				mg/L	12	Standard
Er	166	5427.6	1.3				mg/L	20	Standard
I	127	10600.4	5.7				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.142	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.821	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144302PS WG574573-01
Report Date/Time: Tuesday, June 28, 2016 23:50:11
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.542
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.551
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606144302PS WG574573-01
Report Date/Time: Tuesday, June 28, 2016 23:50:11
Page 3

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606144302SDL WG574573-02
Sample Date/Time: Tuesday, June 28, 2016 23:51:05
Number of Replicates: 3
Autosampler Position: 409
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73539.7	0.5				ug/L	81760	Standard
	Be	9	228.3	17.7	**0.2458**	0.045	18.1	ug/L	17	Standard
	Al	27	1128305.0	2.6	**8.1231**	0.173	2.1	ug/L	1410	Standard
	Sc	45	40995.4	3.1				ug/L	39543	Standard
	Ti	47	3133.3	3.6	**12.6953**	0.331	2.6	ug/L	39	Standard
	V	51	9700.8	3.1	**1.2183**	0.050	4.1	ug/L	1415	Standard
	Cr	52	13743.6	3.1	**0.9582**	0.044	4.6	ug/L	6946	Standard
	Cr	53	1285.1	5.5	**1.1537**	0.099	8.6	ug/L	328	Standard
	Mn	55	302649.7	2.2	**47.5055**	0.538	1.1	ug/L	823	Standard
	Co	59	1856.8	5.3	**0.2384**	0.012	4.9	ug/L	252	Standard
	Ni	60	1619.8	3.4	**0.7729**	0.024	3.1	ug/L	192	Standard
	Cu	65	2766.6	1.1	**1.3956**	0.008	0.6	ug/L	141	Standard
	Zn	66	11859.0	1.8	**10.2169**	0.056	0.5	ug/L	491	Standard
>	**Ge**	72	504593.5	1.7				ug/L	515087	Standard
	As	75	1530.8	2.0	**1.4307**	0.048	3.3	ug/L	20	Standard
	Se	82	33.4	31.2	**0.1774**	0.094	53.0	ug/L	15	Standard
	Se-1	77	64.7	12.6	**0.1577**	0.113	71.7	ug/L	48	Standard
>	**Ga**	71	975.0	10.5				mg/L	90	Standard
	Rb	85	4577.4	4.1				ug/L	23	Standard
	Y	89	452862.0	1.2				ug/L	448510	Standard
>	**Rh**	103	23.3	32.7				ug/L	15	Standard
	Mo	98	722.3	4.5	**0.2491**	0.011	4.3	ug/L	11	Standard
	Ag	107	159.0	11.1	**0.0076**	0.003	35.4	ug/L	134	Standard
	Cd	111	175.3	14.8	**0.0794**	0.012	14.8	mg/L	7	Standard
	Cd	114	462.3	4.9	**0.0844**	0.004	4.5	ug/L	16	Standard
>	**In**	115	608749.5	0.7				ug/L	611117	Standard
	Sn	118	845.0	9.9	**0.0416**	0.014	34.5	ug/L	518	Standard
	Sb	123	2112.7	19.6	**0.4370**	0.091	20.8	ug/L	127	Standard
	Ba	135	21046.6	3.7	**10.2079**	0.317	3.1	ug/L	45	Standard
	Ce	140	96335.3	1.6				ug/L	298	Standard
>	**Tb**	159	932545.8	1.3				ug/L	977942	Standard
	Ho	165	1166.7	10.3				ug/L	12	Standard
	Tl	203	511.3	6.9	**0.0588**	0.004	6.8	ug/L	31	Standard
	Tl	205	456.7	14.9	**0.0631**	0.009	14.8	ug/L	20	Standard
	Pb	206	50001.0	0.8	**9.6268**	0.026	0.3	ug/L	286	Standard
	Pb	207	41687.9	1.1	**8.8985**	0.042	0.5	ug/L	257	Standard
	Pb	208	175842.9	0.5	**9.7051**	0.030	0.3	ug/L	956	Standard
	U	238	977.7	3.5	**0.1400**	0.006	4.1	ug/L	17	Standard
>	**Bi**	209	531859.1	0.7				ug/L	557080	Standard

Sample ID: L1606144302SDL WG574573-02
Report Date/Time: Tuesday, June 28, 2016 23:53:22
Page 1

Approved: June 29, 2016

Na	23	3.3	86.6	**20.9199**	18.126	86.6	mg/L	0	Standard
Mg	24	148.3	20.6	**0.1326**	0.034	25.5	mg/L	32	Standard
K	39	35.0	24.7	**-0.0267**	0.032	120.4	mg/L	30	Standard
Ca	43	75.0	20.0	**5.6602**	2.270	40.1	mg/L	48	Standard
Fe	54	775.4	6.3	**0.5410**	0.062	11.5	mg/L	159	Standard
Fe	57	435.0	3.0	**0.7919**	0.003	0.4	mg/L	195	Standard
> **Sc-1**	45	40995.4	3.1				mg/L	39543	Standard
Cl	35	32746.0	2.5				ug/L	27642	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1743.4	12.5				ug/L	1647	Standard
P	31	17056.3	1.5				ug/L	15231	Standard
S	34	2336.8	6.2				ug/L	1845	Standard
Sr	88	78.3	20.5				ug/L	70	Standard
C	12	126.7	27.7				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	89.2				mg/L	0	Standard
Dy	164	1870.0	3.5				mg/L	12	Standard
Ho-1	165	1166.7	10.3				mg/L	12	Standard
Er	166	1116.7	18.2				mg/L	20	Standard
I	127	3388.7	11.3				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		89.946	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.963	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144302SDL WG574573-02
Report Date/Time: Tuesday, June 28, 2016 23:53:22
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.613
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.473
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: L1606144302SDL WG574573-02
Sample Date/Time: Tuesday, June 28, 2016 23:54:17
Number of Replicates: 3
Autosampler Position: 410
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70935.7	2.2				ug/L	81760	Standard
	Be	9	55.0	15.7	**0.0540**	0.009	16.0	ug/L	17	Standard
	Al	27	218570.5	1.6	**1.6201**	0.011	0.7	ug/L	1410	Standard
	Sc	45	39984.4	1.2				ug/L	39543	Standard
	Ti	47	627.7	4.3	**2.4198**	0.114	4.7	ug/L	39	Standard
	V	51	2847.0	3.0	**0.2129**	0.013	6.1	ug/L	1415	Standard
	Cr	52	7618.6	1.2	**0.0637**	0.011	17.1	ug/L	6946	Standard
	Cr	53	580.0	9.6	**0.3354**	0.068	20.2	ug/L	328	Standard
	Mn	55	61547.4	1.3	**9.5745**	0.188	2.0	ug/L	823	Standard
	Co	59	523.0	3.3	**0.0449**	0.003	6.0	ug/L	252	Standard
	Ni	60	493.7	5.6	**0.1635**	0.016	10.1	ug/L	192	Standard
	Cu	65	675.0	3.2	**0.2729**	0.010	3.6	ug/L	141	Standard
	Zn	66	4565.4	1.3	**3.4380**	0.075	2.2	ug/L	491	Standard
>	**Ge**	72	502988.1	0.7				ug/L	515087	Standard
	As	75	269.2	10.7	**0.2856**	0.028	9.7	ug/L	20	Standard
	Se	82	20.7	22.3	**0.0646**	0.042	64.7	ug/L	15	Standard
	Se-1	77	45.7	15.4	**-0.0984**	0.093	94.5	ug/L	48	Standard
>	**Ga**	71	186.7	12.1				mg/L	90	Standard
	Rb	85	835.0	12.4				ug/L	23	Standard
	Y	89	441201.2	2.7				ug/L	448510	Standard
>	**Rh**	103	13.3	94.4				ug/L	15	Standard
	Mo	98	189.0	7.0	**0.0532**	0.006	12.0	ug/L	11	Standard
	Ag	107	91.3	11.0	**-0.0021**	0.002	87.0	ug/L	134	Standard
	Cd	111	37.0	12.7	**0.0150**	0.002	15.9	mg/L	7	Standard
	Cd	114	85.9	30.8	**0.0122**	0.005	43.3	ug/L	16	Standard
>	**In**	115	591854.8	2.3				ug/L	611117	Standard
	Sn	118	648.3	6.3	**0.0113**	0.007	65.4	ug/L	518	Standard
	Sb	123	522.8	33.2	**0.1022**	0.041	39.7	ug/L	127	Standard
	Ba	135	4107.2	1.3	**2.0294**	0.025	1.2	ug/L	45	Standard
	Ce	140	18988.6	1.0				ug/L	298	Standard
>	**Tb**	159	914076.3	1.3				ug/L	977942	Standard
	Ho	165	253.3	13.1				ug/L	12	Standard
	Tl	203	112.3	10.6	**0.0094**	0.002	17.3	ug/L	31	Standard
	Tl	205	90.0	29.4	**0.0108**	0.004	36.1	ug/L	20	Standard
	Pb	206	10230.5	2.1	**1.9438**	0.049	2.5	ug/L	286	Standard
	Pb	207	8531.4	1.5	**1.7976**	0.010	0.6	ug/L	257	Standard
	Pb	208	36086.5	0.6	**1.9686**	0.029	1.5	ug/L	956	Standard
	U	238	209.0	10.0	**0.0280**	0.003	11.9	ug/L	17	Standard
>	**Bi**	209	525348.3	0.9				ug/L	557080	Standard

Sample ID: L1606144302SDL WG574573-02
Report Date/Time: Tuesday, June 28, 2016 23:56:34
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	68.3	18.4	**0.0329**	0.017	52.2	mg/L	32	Standard
K	39	41.7	42.1	**-0.0025**	0.057	2323.2	mg/L	30	Standard
Ca	43	35.0	49.5	**-0.0873**	2.606	2984.1	mg/L	48	Standard
Fe	54	276.1	10.3	**0.1102**	0.028	25.1	mg/L	159	Standard
Fe	57	273.3	14.9	**0.2889**	0.146	50.5	mg/L	195	Standard
> **Sc-1**	45	39984.4	1.2				mg/L	39543	Standard
Cl	35	32328.5	1.4				ug/L	27642	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	1603.4	8.8				ug/L	1647	Standard
P	31	16148.6	1.7				ug/L	15231	Standard
S	34	2251.8	6.8				ug/L	1845	Standard
Sr	88	76.7	35.9				ug/L	70	Standard
C	12	146.7	10.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	330.4	18.3				mg/L	12	Standard
Ho-1	165	253.3	13.1				mg/L	12	Standard
Er	166	270.0	22.5				mg/L	20	Standard
I	127	1720.1	18.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		86.761	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.651	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144302SDL WG574573-02
Report Date/Time: Tuesday, June 28, 2016 23:56:34
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.848
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.304
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, June 28, 2016 23:57:30
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	72062.8	4.4				ug/L	81760	Standard
	Be	9	45152.2	3.9	**51.5939**	1.461	2.8	ug/L	17	Standard
	Al	27	7383002.3	3.7	**54.3407**	0.756	1.4	ug/L	1410	Standard
	Sc	45	39518.2	3.2				ug/L	39543	Standard
	Ti	47	24690.4	0.9	**102.0722**	0.550	0.5	ug/L	39	Standard
	V	51	333814.0	1.2	**49.2501**	0.769	1.6	ug/L	1415	Standard
	Cr	52	345665.5	1.2	**50.0774**	0.577	1.2	ug/L	6946	Standard
	Cr	53	43291.8	1.9	**50.4406**	0.792	1.6	ug/L	328	Standard
	Mn	55	319563.2	0.6	**50.6110**	0.452	0.9	ug/L	823	Standard
	Co	59	337944.7	0.8	**49.4985**	0.535	1.1	ug/L	252	Standard
	Ni	60	91064.9	0.9	**49.6861**	0.418	0.8	ug/L	192	Standard
	Cu	65	92447.5	0.3	**50.0106**	0.369	0.7	ug/L	141	Standard
	Zn	66	53973.0	1.5	**49.8813**	0.957	1.9	ug/L	491	Standard
>	Ge	72	500217.2	0.4				ug/L	515087	Standard
	As	75	53742.2	1.5	**49.2630**	0.862	1.7	ug/L	20	Standard
	Se	82	5467.6	1.7	**48.7858**	1.017	2.1	ug/L	15	Standard
	Se-1	77	3788.8	1.0	**51.0702**	0.720	1.4	ug/L	48	Standard
>	Ga	71	51.7	5.6				mg/L	90	Standard
	Rb	85	820.0	10.6				ug/L	23	Standard
	Y	89	441206.7	0.2				ug/L	448510	Standard
>	Rh	103	33.3	8.7				ug/L	15	Standard
	Mo	98	251795.6	1.0	**94.4274**	1.474	1.6	ug/L	11	Standard
	Ag	107	307525.6	0.6	**46.4306**	0.349	0.8	ug/L	134	Standard
	Cd	111	100206.1	1.3	**47.5308**	1.177	2.5	mg/L	7	Standard
	Cd	114	244073.0	1.7	**47.6209**	0.821	1.7	ug/L	16	Standard
>	In	115	602825.2	1.4				ug/L	611117	Standard
	Sn	118	281047.8	1.1	**48.0844**	0.236	0.5	ug/L	518	Standard
	Sb	123	222239.4	0.2	**47.7427**	0.588	1.2	ug/L	127	Standard
	Ba	135	96470.5	0.6	**47.3499**	0.512	1.1	ug/L	45	Standard
	Ce	140	125.0	48.0				ug/L	298	Standard
>	Tb	159	921870.8	1.1				ug/L	977942	Standard
	Ho	165	11.7	65.5				ug/L	12	Standard
	Tl	203	392459.6	0.5	**49.4989**	0.479	1.0	ug/L	31	Standard
	Tl	205	352339.7	0.4	**51.1783**	0.770	1.5	ug/L	20	Standard
	Pb	206	250499.8	0.3	**49.1963**	0.428	0.9	ug/L	286	Standard
	Pb	207	226961.6	0.5	**49.4195**	0.337	0.7	ug/L	257	Standard
	Pb	208	926488.8	0.6	**52.1497**	0.850	1.6	ug/L	956	Standard
	U	238	343296.3	1.3	**50.9156**	0.444	0.9	ug/L	17	Standard
>	Bi	209	524204.3	1.1				ug/L	557080	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:59:47
Page 1

Approved: June 29, 2016

Na	23	1.7	173.2	**10.4692**	18.124	173.1	mg/L	0	Standard
Mg	24	3737.1	0.3	**4.9245**	0.180	3.7	mg/L	32	Standard
K	39	1403.4	10.2	**4.6088**	0.536	11.6	mg/L	30	Standard
Ca	43	78.3	16.1	**6.5318**	1.619	24.8	mg/L	48	Standard
Fe	54	5604.3	1.6	**4.9425**	0.217	4.4	mg/L	159	Standard
Fe	57	1713.4	2.9	**5.1658**	0.187	3.6	mg/L	195	Standard
Sc-1	45	39518.2	3.2				mg/L	39543	Standard
Cl	35	33439.5	0.8				ug/L	27642	Standard
Kr	83	2.7	57.3				ug/L	2	Standard
Br	81	1523.4	5.0				ug/L	1647	Standard
P	31	18217.6	1.7				ug/L	15231	Standard
S	34	2641.9	8.4				ug/L	1845	Standard
Sr	88	48.3	11.9				ug/L	70	Standard
C	12	93.3	12.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	0	Standard
Dy	164	9.4	104.5				mg/L	12	Standard
Ho-1	165	11.7	65.5				mg/L	12	Standard
Er	166	13.3	114.6				mg/L	20	Standard
I	127	1510.1	17.2				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	103.188		
Al	27	108.681		
Sc	45			
Ti	47	102.072		
V	51	98.500		
Cr	52	100.155		
Cr	53			
Mn	55	101.222		
Co	59	98.997		
Ni	60	99.372		
Cu	65	100.021		
Zn	66	99.763		
Ge	72		97.113	
As	75	98.526		
Se	82	97.572		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, June 28, 2016 23:59:47
Page 2

Approved: June 29, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	94.427		
Ag	107	92.861		
Cd	111	95.062		
Cd	114			
> In	115			98.643
Sn	118	96.169		
Sb	123	95.485		
Ba	135	94.700		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	98.998		
Tl	205			
Pb	206			
Pb	207			
Pb	208	104.299		
U	238	101.831		
> Bi	209			94.098
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: June 29, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 00:00:41
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70070.0	1.8				ug/L	81760	Standard
	Be	9	30.0	16.7	**0.0256**	0.006	25.2	ug/L	17	Standard
	Al	27	3812.2	30.7	**0.0144**	0.008	57.9	ug/L	1410	Standard
	Sc	45	40014.5	4.3				ug/L	39543	Standard
	Ti	47	26.7	12.1	**-0.0568**	0.013	23.6	ug/L	39	Standard
	V	51	1145.1	6.3	**-0.0403**	0.010	25.4	ug/L	1415	Standard
	Cr	52	6171.9	1.7	**-0.1619**	0.014	8.6	ug/L	6946	Standard
	Cr	53	411.7	3.9	**0.1320**	0.019	14.5	ug/L	328	Standard
	Mn	55	924.7	19.7	**-0.0044**	0.029	653.9	ug/L	823	Standard
	Co	59	261.3	12.2	**0.0062**	0.005	75.0	ug/L	252	Standard
	Ni	60	171.3	8.8	**-0.0131**	0.008	64.0	ug/L	192	Standard
	Cu	65	132.0	5.3	**-0.0207**	0.004	18.6	ug/L	141	Standard
	Zn	66	427.3	9.5	**-0.4339**	0.039	8.9	ug/L	491	Standard
>	**Ge**	72	510161.6	0.4				ug/L	515087	Standard
	As	75	-21.8	127.1	**0.0207**	0.025	120.4	ug/L	20	Standard
	Se	82	17.0	25.9	**0.0293**	0.038	129.6	ug/L	15	Standard
	Se-1	77	53.7	12.0	**0.0005**	0.088	17869.4	ug/L	48	Standard
>	**Ga**	71	35.0	28.6				mg/L	90	Standard
	Rb	85	33.3	60.6				ug/L	23	Standard
	Y	89	437929.1	1.0				ug/L	448510	Standard
>	**Rh**	103	6.7	43.3				ug/L	15	Standard
	Mo	98	269.2	32.9	**0.0838**	0.033	39.3	ug/L	11	Standard
	Ag	107	120.7	16.7	**0.0024**	0.003	124.0	ug/L	134	Standard
	Cd	111	21.8	34.3	**0.0077**	0.004	48.1	mg/L	7	Standard
	Cd	114	36.7	62.4	**0.0024**	0.005	190.6	ug/L	16	Standard
>	**In**	115	590666.4	0.9				ug/L	611117	Standard
	Sn	118	743.4	12.2	**0.0282**	0.017	60.4	ug/L	518	Standard
	Sb	123	532.5	22.4	**0.1040**	0.027	25.7	ug/L	127	Standard
	Ba	135	71.0	29.1	**0.0104**	0.010	96.8	ug/L	45	Standard
	Ce	140	91.7	65.8				ug/L	298	Standard
>	**Tb**	159	911246.7	0.8				ug/L	977942	Standard
	Ho	165	10.0	50.0				ug/L	12	Standard
	Tl	203	81.7	41.2	**0.0055**	0.004	77.1	ug/L	31	Standard
	Tl	205	68.3	66.0	**0.0076**	0.007	85.7	ug/L	20	Standard
	Pb	206	297.7	10.9	**-0.0054**	0.007	123.4	ug/L	286	Standard
	Pb	207	283.7	16.0	**0.0035**	0.010	283.6	ug/L	257	Standard
	Pb	208	1155.4	11.3	**0.0043**	0.008	176.0	ug/L	956	Standard
	U	238	70.3	39.4	**0.0074**	0.004	55.5	ug/L	17	Standard
>	**Bi**	209	527005.8	1.2				ug/L	557080	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 00:02:58
Page 1

Approved: June 29, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	25.0	20.0	**-0.0244**	0.005	21.9	mg/L	32	Standard
K	39	18.3	31.5	**-0.0799**	0.021	26.0	mg/L	30	Standard
Ca	43	16.7	17.3	**-2.8535**	0.318	11.1	mg/L	48	Standard
Fe	54	122.6	39.6	**-0.0261**	0.047	179.6	mg/L	159	Standard
Fe	57	230.0	17.0	**0.1451**	0.148	101.7	mg/L	195	Standard
> **Sc-1**	45	40014.5	4.3				mg/L	39543	Standard
Cl	35	33255.8	1.4				ug/L	27642	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	1523.4	4.7				ug/L	1647	Standard
P	31	17488.4	2.6				ug/L	15231	Standard
S	34	2358.5	3.1				ug/L	1845	Standard
Sr	88	70.0	25.8				ug/L	70	Standard
C	12	73.3	39.4				mg/L	113	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	0	Standard
Dy	164	9.5	100.0				mg/L	12	Standard
Ho-1	165	10.0	50.0				mg/L	12	Standard
Er	166	10.0	100.0				mg/L	20	Standard
I	127	1368.4	18.0				mg/L	1162	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.044	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 00:02:58
Page 2

Approved: June 29, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.654
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		94.601
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 00:02:58
Page 3

Approved: June 29, 2016

MassCal File Name

Mass Calibration File Name Default.tun
MassCal File Path C:\NexIONData\MassCal\Default.tun
Peak Search Window: 1.00

Sample Information

Sample Date/Time: Wednesday, June 29, 2016 09:50:03

Mass Calibration and Resolution

Analyte	E Mass	Meas Mass	Mass C DAC Val	Res DAC Value	Meas Peak W	Custom Res
Li	7.016	6.975	1323	2024	0.699	
Mg	23.985	23.975	4506	2020	0.705	
Co	58.933	58.975	11694	2022	0.697	
In	114.904	114.925	22865	2028	0.690	
U	238.050	238.025	47450	2043	0.688	

Relative Std. Dev.

Mass	Meas. Intens. RSD
5.525	9.254
5.575	1.459
5.625	2.417
5.675	3.442
5.725	2.231
5.775	2.296
5.825	3.784
5.875	3.197
5.925	2.400
5.975	1.087
6.025	1.670
6.075	3.521
6.125	2.277
6.175	2.559
6.225	20.588
6.275	74.689
6.325	108.327
6.375	83.853
6.425	69.722
6.475	32.808
6.525	10.642
6.575	8.985
6.625	1.429
6.675	0.941
6.725	0.857
6.775	3.333
6.825	0.822

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 1

Approved: July 04, 2016

6.875	1.277
6.925	1.899
6.975	1.561
7.025	1.195
7.075	1.059
7.125	1.545
7.175	1.743
7.225	3.248
7.275	8.142
7.325	57.701
7.375	91.287
7.425	51.602
7.475	55.902
7.525	100.000
7.575	69.869
7.625	81.441
7.675	81.441
7.725	70.711
7.775	55.902
7.825	59.266
7.875	49.793
7.925	104.583
7.975	98.543
8.025	94.786
8.075	223.607
8.125	81.441
8.175	84.265
8.225	61.237
8.275	81.312
8.325	81.441
8.375	91.287
8.425	82.402
8.475	22.822
22.525	149.071
22.575	52.705
22.625	46.771
22.675	48.627
22.725	39.087
22.775	18.109
22.825	22.048
22.875	36.422
22.925	32.723
22.975	82.685
23.025	39.239
23.075	25.529
23.125	43.064
23.175	21.317

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 2

Approved: July 04, 2016

23.225	22.232
23.275	15.408
23.325	42.127
23.375	37.172
23.425	24.296
23.475	14.668
23.525	3.807
23.575	1.593
23.625	2.258
23.675	1.180
23.725	1.542
23.775	1.943
23.825	1.185
23.875	1.595
23.925	1.324
23.975	1.055
24.025	1.809
24.075	1.103
24.125	1.249
24.175	0.652
24.225	1.203
24.275	3.614
24.325	39.316
24.375	52.355
24.425	20.738
24.475	5.001
24.525	5.067
24.575	0.930
24.625	1.241
24.675	0.972
24.725	1.552
24.775	0.581
24.825	0.736
24.875	1.684
24.925	0.952
24.975	1.395
25.025	0.762
25.075	1.498
25.125	1.418
25.175	1.145
25.225	6.288
25.275	44.605
25.325	50.990
25.375	73.823
25.425	57.334
25.475	7.010
57.525	8.193

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 3

Approved: July 04, 2016

57.575	3.609
57.625	2.367
57.675	3.747
57.725	3.869
57.775	3.226
57.825	1.406
57.875	2.220
57.925	2.049
57.975	2.397
58.025	3.045
58.075	0.897
58.125	1.397
58.175	1.608
58.225	1.481
58.275	5.244
58.325	6.996
58.375	14.816
58.425	39.994
58.475	18.390
58.525	7.862
58.575	1.876
58.625	1.684
58.675	0.757
58.725	2.431
58.775	2.139
58.825	2.293
58.875	1.987
58.925	2.134
58.975	1.760
59.025	2.893
59.075	1.855
59.125	4.126
59.175	3.545
59.225	2.980
59.275	6.918
59.325	57.198
59.375	79.057
59.425	39.123
59.475	32.983
59.525	7.337
59.575	10.704
59.625	4.888
59.675	5.230
59.725	3.575
59.775	2.279
59.825	3.495
59.875	0.859

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 4

Approved: July 04, 2016

59.925	2.164
59.975	4.187
60.025	3.749
60.075	1.767
60.125	4.373
60.175	1.581
60.225	1.647
60.275	29.768
60.325	42.377
60.375	71.261
60.425	71.261
60.475	69.722
113.525	10.800
113.575	6.260
113.625	1.672
113.675	1.646
113.725	1.420
113.775	2.281
113.825	1.446
113.875	0.820
113.925	1.607
113.975	1.407
114.025	1.046
114.075	2.403
114.125	2.090
114.175	3.331
114.225	5.316
114.275	12.354
114.325	37.872
114.375	26.082
114.425	15.800
114.475	3.831
114.525	4.062
114.575	1.808
114.625	2.980
114.675	3.272
114.725	2.895
114.775	1.904
114.825	1.941
114.875	1.809
114.925	0.906
114.975	0.414
115.025	2.617
115.075	0.808
115.125	1.618
115.175	2.428
115.225	3.339

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 5

Approved: July 04, 2016

115.275	5.459
115.325	14.203
115.375	69.869
115.425	66.254
115.475	26.966
115.525	18.235
115.575	7.589
115.625	3.788
115.675	2.661
115.725	3.661
115.775	2.867
115.825	3.952
115.875	6.794
115.925	4.558
115.975	5.421
116.025	1.319
116.075	4.928
116.125	7.259
116.175	4.689
116.225	9.289
116.275	22.491
116.325	26.307
116.375	26.146
116.425	67.711
116.475	40.825
236.525	
236.575	13.258
236.625	39.219
236.675	66.333
236.725	37.815
236.775	20.039
236.825	39.510
236.875	22.934
236.925	27.407
236.975	8.537
237.025	28.853
237.075	45.928
237.125	41.937
237.175	36.377
237.225	26.058
237.275	27.285
237.325	43.196
237.375	45.018
237.425	23.111
237.475	20.908
237.525	26.413
237.575	10.671

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 6

Approved: July 04, 2016

237.625	6.558
237.675	4.584
237.725	2.707
237.775	1.019
237.825	0.867
237.875	0.837
237.925	1.315
237.975	0.556
238.025	1.099
238.075	0.473
238.125	0.980
238.175	0.943
238.225	1.564
238.275	1.714
238.325	1.785
238.375	2.057
238.425	2.571
238.475	6.163
238.525	10.798
238.575	23.407
238.625	35.493
238.675	33.171
238.725	15.721
238.775	9.421
238.825	23.516
238.875	27.721
238.925	14.417
238.975	30.010
239.025	24.993
239.075	29.933
239.125	17.678
239.175	38.433
239.225	46.981
239.275	15.215
239.325	48.601
239.375	50.659
239.425	23.570
239.475	38.955

Approved: July 04, 2016

SmartTune Wizard - Summary

```
Optimization Summary

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Start Time: 6/29/2016 10:01:57 AM
End Time: 6/29/2016 10:04:08 AM

Daily Performance Check - [Passed]  Optimum value(s): N/A
        Obtained Intensity (Be 9.0122): 4651.78
        Obtained Intensity (Mg 23.985): 162439.88
        Obtained Intensity (In 114.904): 66202.13
        Obtained Intensity (U 238.05): 140216.56
        Obtained Intensity (Bkgd 220): 0.33
        Obtained Formula (CeO 155.9 / Ce 139.905): 0.018 (=3640.50 / 205153.73)
        Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.004 (=909.89 / 205153.73)
```

Report Date/Time: Wednesday, June 29, 2016 10:04:08
Page 1

Approved: July 04, 2016

SmartTune Wizard - Details

```
Optimization Details

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Optimization Status

Start Time: 6/29/2016 10:01:57 AM

Daily Performance Check
        Optimization Settings:
                Method: C:\NexIONData\Method\ESI Daily Performance.mth.
                Intensity Criterion: Be 9.0122 > 2000
                Intensity Criterion: Mg 23.985 > 15000
                Intensity Criterion: In 114.904 > 40000
                Intensity Criterion: U 238.05 > 30000
                Intensity Criterion: Bkgd 220 <= 5
                Formula Criterion: CeO 155.9 / Ce 139.905 <= 0.025
                Formula Criterion: Ce++ 69.9527 / Ce 139.905 <= 0.03

        Optimization Results:
        Initial Try
                Obtained Intensity (Be 9.0122): 4651.78
                Obtained Intensity (Mg 23.985): 162439.88
                Obtained Intensity (In 114.904): 66202.13
                Obtained Intensity (U 238.05): 140216.56
                Obtained Intensity (Bkgd 220): 0.33
                Obtained Formula (CeO 155.9 / Ce 139.905): 0.018 (=3640.50 / 205153.73)
                Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.004 (=909.89 / 205153.73)
[Passed]  Optimum value(s): N/A

End Time: 6/29/2016 10:04:08 AM
```


Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Blank
Sample Date/Time: Wednesday, June 29, 2016 11:22:34
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	74698.8	3.1				ug/L		Standard
	Be	9	8.3	124.9				ug/L		Standard
	Al	27	1493.4	9.1				ug/L		Standard
	Sc	45	61029.4	2.9				ug/L		Standard
	Ti	47	35.0	12.5				ug/L		Standard
	V	51	1434.8	4.3				ug/L		Standard
	Cr	52	7634.6	2.5				ug/L		Standard
	Cr	53	363.3	0.8				ug/L		Standard
	Mn	55	747.0	1.6				ug/L		Standard
	Co	59	233.7	7.1				ug/L		Standard
	Ni	60	214.7	3.5				ug/L		Standard
	Cu	65	104.0	12.0				ug/L		Standard
	Zn	66	363.0	3.8				ug/L		Standard
>	Ge	72	495142.3	1.3				ug/L		Standard
	As	75	-49.8	7.8				ug/L		Standard
	Se	82	13.7	23.9				ug/L		Standard
	Se-1	77	42.3	9.8				ug/L		Standard
>	Ga	71	41.7	25.0				mg/L		Standard
	Rb	85	16.7	91.7				ug/L		Standard
	Y	89	417098.4	0.9				ug/L		Standard
>	Rh	103	13.3	86.6				ug/L		Standard
	Mo	98	10.7	10.9				ug/L		Standard
	Ag	107	68.0	6.4				ug/L		Standard
	Cd	111	2.3	24.9				mg/L		Standard
	Cd	114	11.7	106.5				ug/L		Standard
>	In	115	587441.1	0.5				ug/L		Standard
	Sn	118	496.7	9.7				ug/L		Standard
	Sb	123	77.7	34.7				ug/L		Standard
	Ba	135	24.3	19.4				ug/L		Standard
	Ce	140	25.0	20.0				ug/L		Standard
>	Tb	159	939701.0	1.8				ug/L		Standard
	Ho	165	10.0	50.0				ug/L		Standard
	Tl	203	4.0	25.0				ug/L		Standard
	Tl	205	1.7	173.2				ug/L		Standard
	Pb	206	239.7	3.6				ug/L		Standard
	Pb	207	197.7	3.9				ug/L		Standard
	Pb	208	883.3	3.3				ug/L		Standard
	U	238	3.3	17.3				ug/L		Standard
>	Bi	209	523984.5	1.3				ug/L		Standard

Sample ID: Blank
Report Date/Time: Wednesday, June 29, 2016 11:24:51
Page 1

Approved: July 04, 2016

	Na	23	0.0		mg/L	Standard
	Mg	24	36.7	55.1	mg/L	Standard
	K	39	38.3	52.7	mg/L	Standard
	Ca	43	10.0	50.0	mg/L	Standard
	Fe	54	179.1	14.9	mg/L	Standard
	Fe	57	335.0	10.3	mg/L	Standard
>	**Sc-1**	45	61029.4	2.9	mg/L	Standard
	Cl	35	20194.8	0.8	ug/L	Standard
	Kr	83	4.3	70.5	ug/L	Standard
	Br	81	1546.7	2.4	ug/L	Standard
	P	31	15286.0	3.7	ug/L	Standard
	S	34	1571.7	2.7	ug/L	Standard
	Sr	88	73.3	3.9	ug/L	Standard
	C	12	63.3	24.1	mg/L	Standard
	N	14	0.0		mg/L	Standard
	Hg	202	3.3	173.2	mg/L	Standard
	Dy	164	6.0	187.1	mg/L	Standard
	Ho-1	165	10.0	50.0	mg/L	Standard
	Er	166	13.3	86.6	mg/L	Standard
	I	127	1391.7	17.5	mg/L	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Blank
Report Date/Time: Wednesday, June 29, 2016 11:24:51
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Blank
Report Date/Time: Wednesday, June 29, 2016 11:24:51
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 1
Sample Date/Time: Wednesday, June 29, 2016 11:25:45
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75658.5	2.0				ug/L	74699	Standard
	Be	9	1.7	173.2				ug/L	8	Standard
	Al	27	1555.1	3.1				ug/L	1493	Standard
	Sc	45	60145.9	2.9				ug/L	61029	Standard
	Ti	47	32.7	4.7				ug/L	35	Standard
	V	51	1434.9	4.0				ug/L	1435	Standard
	Cr	52	7544.5	2.5				ug/L	7635	Standard
	Cr	53	365.0	18.0				ug/L	363	Standard
	Mn	55	783.7	0.8				ug/L	747	Standard
	Co	59	239.0	2.9				ug/L	234	Standard
	Ni	60	220.7	1.8				ug/L	215	Standard
	Cu	65	103.0	8.3				ug/L	104	Standard
	Zn	66	365.0	4.1				ug/L	363	Standard
>	**Ge**	72	495351.7	2.1				ug/L	495142	Standard
	As	75	-40.0	8.7				ug/L	-50	Standard
	Se	82	19.1	38.4				ug/L	14	Standard
	Se-1	77	47.3	15.6				ug/L	42	Standard
>	**Ga**	71	53.3	47.2				mg/L	42	Standard
	Rb	85	15.0	88.2				ug/L	17	Standard
	Y	89	427717.4	1.6				ug/L	417098	Standard
>	**Rh**	103	8.3	34.6				ug/L	13	Standard
	Mo	98	11.1	31.8				ug/L	11	Standard
	Ag	107	86.3	19.8				ug/L	68	Standard
	Cd	111	1.3	43.8				mg/L	2	Standard
	Cd	114	3.6	152.7				ug/L	12	Standard
>	**In**	115	599155.0	1.0				ug/L	587441	Standard
	Sn	118	486.7	3.3				ug/L	497	Standard
	Sb	123	28.2	25.8				ug/L	78	Standard
	Ba	135	25.3	13.9				ug/L	24	Standard
	Ce	140	31.7	59.8				ug/L	25	Standard
>	**Tb**	159	985563.2	0.5				ug/L	939701	Standard
	Ho	165	6.7	86.6				ug/L	10	Standard
	Tl	203	1.7	91.7				ug/L	4	Standard
	Tl	205	0.0					ug/L	2	Standard
	Pb	206	249.7	4.0				ug/L	240	Standard
	Pb	207	216.0	2.0				ug/L	198	Standard
	Pb	208	873.0	4.1				ug/L	883	Standard
	U	238	3.3	34.6				ug/L	3	Standard
>	**Bi**	209	535312.9	0.6				ug/L	523984	Standard

Sample ID: Standard 1
Report Date/Time: Wednesday, June 29, 2016 11:28:02
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	mg/L	0	Standard
Mg	24	31.7	18.2	mg/L	37	Standard
K	39	26.7	28.6	mg/L	38	Standard
Ca	43	35.0	24.7	mg/L	10	Standard
Fe	54	130.8	15.0	mg/L	179	Standard
Fe	57	305.0	2.8	mg/L	335	Standard
Sc-1	45	60145.9	2.9	mg/L	61029	Standard
Cl	35	20341.6	1.8	ug/L	20195	Standard
Kr	83	3.0	33.3	ug/L	4	Standard
Br	81	1546.7	10.4	ug/L	1547	Standard
P	31	15145.9	1.8	ug/L	15286	Standard
S	34	1610.1	6.1	ug/L	1572	Standard
Sr	88	86.7	14.5	ug/L	73	Standard
C	12	60.0	28.9	mg/L	63	Standard
N	14	0.0		mg/L	0	Standard
Hg	202	0.0		mg/L	3	Standard
Dy	164	5.7	108.5	mg/L	6	Standard
Ho-1	165	6.7	86.6	mg/L	10	Standard
Er	166	20.0	50.0	mg/L	13	Standard
I	127	1368.4	7.8	mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72			
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: Standard 1
Report Date/Time: Wednesday, June 29, 2016 11:28:02
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 1
Report Date/Time: Wednesday, June 29, 2016 11:28:02
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 2
Sample Date/Time: Wednesday, June 29, 2016 11:28:57
Number of Replicates: 3
Autosampler Position: 2
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73081.0	3.5				ug/L	74699	Standard
	Be	9	36.7	56.8				ug/L	8	Standard
	Al	27	9102.7	2.7				ug/L	1493	Standard
	Sc	45	58226.8	3.4				ug/L	61029	Standard
	Ti	47	79.0	2.2				ug/L	35	Standard
	V	51	1848.7	2.3				ug/L	1435	Standard
	Cr	52	7933.1	1.4				ug/L	7635	Standard
	Cr	53	350.0	15.1				ug/L	363	Standard
	Mn	55	1375.4	4.6				ug/L	747	Standard
	Co	59	574.7	5.8				ug/L	234	Standard
	Ni	60	324.3	7.6				ug/L	215	Standard
	Cu	65	285.0	4.9				ug/L	104	Standard
	Zn	66	549.3	0.4				ug/L	363	Standard
>	**Ge**	72	493663.5	1.5				ug/L	495142	Standard
	As	75	-10.4	282.1				ug/L	-50	Standard
	Se	82	20.6	18.5				ug/L	14	Standard
	Se-1	77	49.0	19.5				ug/L	42	Standard
>	**Ga**	71	58.3	27.6				mg/L	42	Standard
	Rb	85	25.0	60.0				ug/L	17	Standard
	Y	89	437555.0	0.8				ug/L	417098	Standard
>	**Rh**	103	25.0	34.6				ug/L	13	Standard
	Mo	98	283.4	4.8				ug/L	11	Standard
	Ag	107	412.3	3.2				ug/L	68	Standard
	Cd	111	121.8	9.1				mg/L	2	Standard
	Cd	114	268.9	5.0				ug/L	12	Standard
>	**In**	115	590951.3	0.9				ug/L	587441	Standard
	Sn	118	850.0	6.4				ug/L	497	Standard
	Sb	123	254.5	2.7				ug/L	78	Standard
	Ba	135	136.0	15.7				ug/L	24	Standard
	Ce	140	18.3	56.8				ug/L	25	Standard
>	**Tb**	159	941723.8	1.9				ug/L	939701	Standard
	Ho	165	3.3	86.6				ug/L	10	Standard
	Tl	203	417.3	5.5				ug/L	4	Standard
	Tl	205	348.3	7.4				ug/L	2	Standard
	Pb	206	633.0	0.8				ug/L	240	Standard
	Pb	207	557.0	2.2				ug/L	198	Standard
	Pb	208	2275.4	1.7				ug/L	883	Standard
	U	238	366.3	0.7				ug/L	3	Standard
>	**Bi**	209	512414.5	1.5				ug/L	523984	Standard

Sample ID: Standard 2
Report Date/Time: Wednesday, June 29, 2016 11:31:14
Page 1

Approved: July 04, 2016

	Na	23	0.0		mg/L	0	Standard
	Mg	24	51.7	24.4	mg/L	37	Standard
	K	39	51.7	24.4	mg/L	38	Standard
	Ca	43	38.3	19.9	mg/L	10	Standard
	Fe	54	160.9	17.5	mg/L	179	Standard
	Fe	57	303.3	9.1	mg/L	335	Standard
>	**Sc-1**	45	58226.8	3.4	mg/L	61029	Standard
	Cl	35	20887.7	1.1	ug/L	20195	Standard
	Kr	83	2.3	107.9	ug/L	4	Standard
	Br	81	1533.4	9.8	ug/L	1547	Standard
	P	31	14952.4	2.3	ug/L	15286	Standard
	S	34	1701.8	4.5	ug/L	1572	Standard
	Sr	88	66.7	31.2	ug/L	73	Standard
	C	12	46.7	65.5	mg/L	63	Standard
	N	14	0.0		mg/L	0	Standard
	Hg	202	0.0		mg/L	3	Standard
	Dy	164	12.7	89.9	mg/L	6	Standard
	Ho-1	165	3.3	86.6	mg/L	10	Standard
	Er	166	13.3	43.3	mg/L	13	Standard
	I	127	1343.4	21.4	mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Standard 2
Report Date/Time: Wednesday, June 29, 2016 11:31:14
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 2
Report Date/Time: Wednesday, June 29, 2016 11:31:14
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 3
Sample Date/Time: Wednesday, June 29, 2016 11:32:08
Number of Replicates: 3
Autosampler Position: 3
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70836.9	0.7				ug/L	74699	Standard
	Be	9	41495.2	4.9	**50.0000**	2.291	4.6	ug/L	8	Standard
	Al	27	6913642.3	1.4	**50.0000**	0.333	0.7	ug/L	1493	Standard
	Sc	45	58109.6	1.6				ug/L	61029	Standard
	Ti	47	38024.7	0.9	**100.0000**	1.896	1.9	ug/L	35	Standard
	V	51	385028.2	1.3	**50.0000**	1.023	2.0	ug/L	1435	Standard
	Cr	52	378864.5	0.8	**50.0000**	0.813	1.6	ug/L	7635	Standard
	Cr	53	46780.4	0.7	**50.0000**	0.853	1.7	ug/L	363	Standard
	Mn	55	354877.0	1.3	**50.0000**	1.137	2.3	ug/L	747	Standard
	Co	59	336932.8	1.2	**50.0000**	1.027	2.1	ug/L	234	Standard
	Ni	60	94633.6	0.8	**50.0000**	0.638	1.3	ug/L	215	Standard
	Cu	65	94252.3	1.0	**50.0000**	0.913	1.8	ug/L	104	Standard
	Zn	66	55983.3	1.5	**50.0000**	1.110	2.2	ug/L	363	Standard
>	**Ge**	72	497274.4	1.0				ug/L	495142	Standard
	As	75	56952.3	0.9	**50.0000**	0.907	1.8	ug/L	-50	Standard
	Se	82	5614.6	2.5	**50.0000**	1.630	3.3	ug/L	14	Standard
	Se-1	77	3760.8	2.5	**50.0000**	1.757	3.5	ug/L	42	Standard
>	**Ga**	71	60.0	22.0				mg/L	42	Standard
	Rb	85	735.0	7.4				ug/L	17	Standard
	Y	89	413691.0	2.8				ug/L	417098	Standard
>	**Rh**	103	28.3	27.0				ug/L	13	Standard
	Mo	98	261684.0	1.3	**100.0000**	1.411	1.4	ug/L	11	Standard
	Ag	107	317580.7	1.1	**50.0000**	0.030	0.1	ug/L	68	Standard
	Cd	111	102024.7	1.3	**50.0000**	0.465	0.9	mg/L	2	Standard
	Cd	114	251691.0	1.5	**50.0000**	1.089	2.2	ug/L	12	Standard
>	**In**	115	567339.3	1.1				ug/L	587441	Standard
	Sn	118	284919.2	1.1	**50.0000**	0.694	1.4	ug/L	497	Standard
	Sb	123	226852.6	0.2	**50.0000**	0.610	1.2	ug/L	78	Standard
	Ba	135	99575.8	0.1	**50.0000**	0.595	1.2	ug/L	24	Standard
	Ce	140	83.3	34.6				ug/L	25	Standard
>	**Tb**	159	932165.5	0.9				ug/L	939701	Standard
	Ho	165	13.3	21.7				ug/L	10	Standard
	Tl	203	405993.6	0.4	**50.0000**	0.749	1.5	ug/L	4	Standard
	Tl	205	349857.0	1.5	**50.0000**	0.282	0.6	ug/L	2	Standard
	Pb	206	265201.9	0.6	**50.0000**	0.817	1.6	ug/L	240	Standard
	Pb	207	235754.7	0.6	**50.0000**	0.829	1.7	ug/L	198	Standard
	Pb	208	926203.5	0.6	**50.0000**	0.626	1.3	ug/L	883	Standard
	U	238	337413.2	1.0	**50.0000**	0.543	1.1	ug/L	3	Standard
>	**Bi**	209	511781.8	1.3				ug/L	523984	Standard

Sample ID: Standard 3
Report Date/Time: Wednesday, June 29, 2016 11:34:25
Page 1

Approved: July 04, 2016

Na	23	0.0					mg/L	0	Standard
Mg	24	3782.1	6.2	**5.0000**	0.388	7.8	mg/L	37	Standard
K	39	2271.8	5.5	**5.0000**	0.254	5.1	mg/L	38	Standard
Ca	43	120.0	23.2	**5.0000**	1.671	33.4	mg/L	10	Standard
Fe	54	5877.0	6.9	**5.0000**	0.418	8.4	mg/L	179	Standard
Fe	57	1703.4	4.4	**5.0000**	0.228	4.6	mg/L	335	Standard
> **Sc-1**	45	58109.6	1.6				mg/L	61029	Standard
Cl	35	22061.4	1.9				ug/L	20195	Standard
Kr	83	2.7	94.4				ug/L	4	Standard
Br	81	1433.4	2.5				ug/L	1547	Standard
P	31	17349.9	2.8				ug/L	15286	Standard
S	34	2026.8	6.8				ug/L	1572	Standard
Sr	88	76.7	26.4				ug/L	73	Standard
C	12	43.3	48.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	33.0	63.6				mg/L	6	Standard
Ho-1	165	13.3	21.7				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1500.1	14.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72			
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: Standard 3
Report Date/Time: Wednesday, June 29, 2016 11:34:25
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 3
Report Date/Time: Wednesday, June 29, 2016 11:34:25
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 4
Sample Date/Time: Wednesday, June 29, 2016 11:35:20
Number of Replicates: 3
Autosampler Position: 4
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	70562.3	0.6				ug/L	74699	Standard
	Be	9	83650.3	2.2	**100.5872**	1.689	1.7	ug/L	8	Standard
	Al	27	13996690.0	1.6	**100.8115**	1.548	1.5	ug/L	1493	Standard
	Sc	45	58977.9	1.4				ug/L	61029	Standard
	Ti	47	75710.7	1.4	**200.4046**	3.310	1.7	ug/L	35	Standard
	V	51	745374.0	1.7	**98.8583**	0.465	0.5	ug/L	1435	Standard
	Cr	52	747654.3	1.6	**100.2327**	0.593	0.6	ug/L	7635	Standard
	Cr	53	92243.3	0.8	**99.8509**	0.489	0.5	ug/L	363	Standard
	Mn	55	686210.3	1.7	**98.7774**	1.044	1.1	ug/L	747	Standard
	Co	59	647972.6	0.9	**98.4550**	0.479	0.5	ug/L	234	Standard
	Ni	60	186003.6	1.6	**99.5867**	0.367	0.4	ug/L	215	Standard
	Cu	65	184421.8	1.6	**99.3502**	0.739	0.7	ug/L	104	Standard
	Zn	66	108058.7	1.7	**98.8348**	0.723	0.7	ug/L	363	Standard
>	Ge	72	493312.7	1.3				ug/L	495142	Standard
	As	75	109996.7	2.4	**98.6160**	1.152	1.2	ug/L	-50	Standard
	Se	82	10772.1	0.2	**98.3851**	1.295	1.3	ug/L	14	Standard
	Se-1	77	7226.7	2.3	**98.6814**	1.085	1.1	ug/L	42	Standard
>	Ga	71	88.3	19.9				mg/L	42	Standard
	Rb	85	1528.4	8.2				ug/L	17	Standard
	Y	89	411133.1	2.0				ug/L	417098	Standard
>	Rh	103	50.0	20.0				ug/L	13	Standard
	Mo	98	506691.6	2.0	**196.6866**	6.239	3.2	ug/L	11	Standard
	Ag	107	624459.9	0.9	**99.1146**	1.822	1.8	ug/L	68	Standard
	Cd	111	200021.3	0.6	**98.9639**	1.657	1.7	mg/L	2	Standard
	Cd	114	487602.9	1.3	**98.3617**	2.136	2.2	ug/L	12	Standard
>	In	115	567890.6	1.1				ug/L	587441	Standard
	Sn	118	560558.1	0.7	**99.1782**	1.388	1.4	ug/L	497	Standard
	Sb	123	448476.6	1.3	**99.3779**	2.340	2.4	ug/L	78	Standard
	Ba	135	196432.8	0.6	**99.2744**	1.775	1.8	ug/L	24	Standard
	Ce	140	161.7	46.3				ug/L	25	Standard
>	Tb	159	938936.1	1.2				ug/L	939701	Standard
	Ho	165	10.0					ug/L	10	Standard
	Tl	203	803908.2	0.6	**99.1382**	1.002	1.0	ug/L	4	Standard
	Tl	205	702848.2	0.9	**99.8673**	1.348	1.3	ug/L	2	Standard
	Pb	206	517801.9	1.0	**98.4705**	1.644	1.7	ug/L	240	Standard
	Pb	207	459764.9	1.8	**98.4124**	2.372	2.4	ug/L	198	Standard
	Pb	208	1827348.6	0.4	**98.9940**	0.966	1.0	ug/L	883	Standard
	U	238	674609.9	0.9	**99.6270**	1.321	1.3	ug/L	3	Standard
>	Bi	209	515462.4	0.6				ug/L	523984	Standard

Sample ID: Standard 4
Report Date/Time: Wednesday, June 29, 2016 11:37:37
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.9850**	34.606	173.2	mg/L	0	Standard
Mg	24	7435.2	2.2	**9.8676**	0.257	2.6	mg/L	37	Standard
K	39	4282.3	4.3	**9.6823**	0.481	5.0	mg/L	38	Standard
Ca	43	178.3	18.2	**9.1297**	2.109	23.1	mg/L	10	Standard
Fe	54	11328.6	1.8	**9.8033**	0.307	3.1	mg/L	179	Standard
Fe	57	3145.3	5.1	**9.9880**	0.532	5.3	mg/L	335	Standard
Sc-1	45	58977.9	1.4				mg/L	61029	Standard
Cl	35	21244.2	2.5				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	1583.4	3.9				ug/L	1547	Standard
P	31	17550.2	2.6				ug/L	15286	Standard
S	34	2090.1	3.5				ug/L	1572	Standard
Sr	88	83.3	42.6				ug/L	73	Standard
C	12	56.7	62.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	22.9	66.8				mg/L	6	Standard
Ho-1	165	10.0					mg/L	10	Standard
Er	166	10.0					mg/L	13	Standard
I	127	1638.4	19.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72			
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: Standard 4
Report Date/Time: Wednesday, June 29, 2016 11:37:37
Page 2

Approved: July 04, 2016

	Rb	85
	Y	89
>	Rh	103
	Mo	98
	Ag	107
	Cd	111
	Cd	114
>	In	115
	Sn	118
	Sb	123
	Ba	135
	Ce	140
>	Tb	159
	Ho	165
	Tl	203
	Tl	205
	Pb	206
	Pb	207
	Pb	208
	U	238
>	Bi	209
	Na	23
	Mg	24
	K	39
	Ca	43
	Fe	54
	Fe	57
>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Corr. Coef.	Na	23	Correlation coefficient < 0.998
Corr. Coef.	Ca	43	Correlation coefficient < 0.998

Sample ID: Standard 4
Report Date/Time: Wednesday, June 29, 2016 11:37:37
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 1
Sample Date/Time: Wednesday, June 29, 2016 11:38:33
Number of Replicates: 3
Autosampler Position: 201
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	69383.6	2.2				ug/L	74699	Standard
	Be	9	41665.5	1.0	**50.9761**	1.262	2.5	ug/L	8	Standard
	Al	27	7007266.1	0.7	**51.3351**	1.064	2.1	ug/L	1493	Standard
	Sc	45	57040.5	0.5				ug/L	61029	Standard
	Ti	47	39782.2	0.5	**104.1738**	1.427	1.4	ug/L	35	Standard
	V	51	390724.1	1.4	**51.2089**	1.334	2.6	ug/L	1435	Standard
	Cr	52	380689.4	0.9	**50.0149**	1.033	2.1	ug/L	7635	Standard
	Cr	53	46596.6	2.0	**49.7615**	1.193	2.4	ug/L	363	Standard
	Mn	55	364452.5	1.9	**51.8606**	1.232	2.4	ug/L	747	Standard
	Co	59	335357.4	2.0	**50.4145**	0.943	1.9	ug/L	234	Standard
	Ni	60	95528.2	0.6	**50.5685**	0.768	1.5	ug/L	215	Standard
	Cu	65	95355.8	0.8	**50.7995**	0.891	1.8	ug/L	104	Standard
	Zn	66	57795.4	0.9	**52.1095**	0.303	0.6	ug/L	363	Standard
>	Ge	72	498415.5	1.2				ug/L	495142	Standard
	As	75	57290.2	0.4	**50.8755**	0.674	1.3	ug/L	-50	Standard
	Se	82	5486.5	1.1	**49.5278**	0.838	1.7	ug/L	14	Standard
	Se-1	77	3797.8	2.0	**51.0456**	1.569	3.1	ug/L	42	Standard
>	Ga	71	88.3	43.2				mg/L	42	Standard
	Rb	85	576.7	2.8				ug/L	17	Standard
	Y	89	413113.3	1.2				ug/L	417098	Standard
>	Rh	103	26.7	92.5				ug/L	13	Standard
	Mo	98	275236.9	1.8	**107.0386**	1.474	1.4	ug/L	11	Standard
	Ag	107	320241.4	0.9	**50.9276**	0.747	1.5	ug/L	68	Standard
	Cd	111	101502.5	0.4	**50.3185**	0.332	0.7	mg/L	2	Standard
	Cd	114	248344.9	2.5	**50.2016**	1.511	3.0	ug/L	12	Standard
>	In	115	566676.1	0.5				ug/L	587441	Standard
	Sn	118	310270.4	1.9	**54.9671**	1.097	2.0	ug/L	497	Standard
	Sb	123	224427.6	0.4	**49.8273**	0.118	0.2	ug/L	78	Standard
	Ba	135	98607.5	0.6	**49.9289**	0.574	1.1	ug/L	24	Standard
	Ce	140	138.3	11.0				ug/L	25	Standard
>	Tb	159	931841.4	1.3				ug/L	939701	Standard
	Ho	165	10.0	86.6				ug/L	10	Standard
	Tl	203	410286.1	0.9	**50.6578**	0.710	1.4	ug/L	4	Standard
	Tl	205	350970.1	0.3	**49.9294**	0.473	0.9	ug/L	2	Standard
	Pb	206	272426.4	0.6	**51.8371**	0.812	1.6	ug/L	240	Standard
	Pb	207	234103.7	0.5	**50.1331**	0.525	1.0	ug/L	198	Standard
	Pb	208	940824.8	0.3	**50.9948**	0.649	1.3	ug/L	883	Standard
	U	238	333596.8	0.3	**49.3228**	0.442	0.9	ug/L	3	Standard
>	Bi	209	514849.0	1.0				ug/L	523984	Standard

Sample ID: QC Std 1
Report Date/Time: Wednesday, June 29, 2016 11:40:50
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	3822.1	2.3	**5.2142**	0.148	2.8	mg/L	37	Standard
K	39	2126.8	6.1	**4.9163**	0.328	6.7	mg/L	38	Standard
Ca	43	118.3	24.4	**5.4681**	1.919	35.1	mg/L	10	Standard
Fe	54	5948.1	0.7	**5.2575**	0.061	1.2	mg/L	179	Standard
Fe	57	1791.8	1.9	**5.4420**	0.142	2.6	mg/L	335	Standard
> **Sc-1**	45	57040.5	0.5				mg/L	61029	Standard
Cl	35	20895.1	0.7				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	1520.1	6.3				ug/L	1547	Standard
P	31	17665.3	1.6				ug/L	15286	Standard
S	34	2033.5	1.7				ug/L	1572	Standard
Sr	88	80.0	10.8				ug/L	73	Standard
C	12	70.0	24.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	23.0	24.5				mg/L	6	Standard
Ho-1	165	10.0	86.6				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1353.4	28.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	101.952		
Al	27	102.670		
Sc	45			
Ti	47	104.174		
V	51	102.418		
Cr	52	100.030		
Cr	53			
Mn	55	103.721		
Co	59	100.829		
Ni	60	101.137		
Cu	65	101.599		
Zn	66	104.219		
> Ge	72		100.661	
As	75	101.751		
Se	82	99.056		
Se-1	77			
> Ga	71			

Sample ID: QC Std 1
Report Date/Time: Wednesday, June 29, 2016 11:40:50
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	107.039		
Ag	107	101.855		
Cd	111	100.637		
Cd	114			
> In	115		96.465	
Sn	118	109.934		
Sb	123	99.655		
Ba	135	99.858		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.316		
Tl	205			
Pb	206	103.674		
Pb	207	100.266		
Pb	208	101.990		
U	238	98.646		
> Bi	209		98.257	
Na	23	0.100		
Mg	24	104.283		
K	39	98.326		
Ca	43	109.361		
Fe	54	105.151		
Fe	57	108.840		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 1	Na	23	
QC Std 1	Ca	43	

Sample ID: QC Std 1
Report Date/Time: Wednesday, June 29, 2016 11:40:50
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 2
Sample Date/Time: Wednesday, June 29, 2016 11:41:46
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70378.1	2.5				ug/L	74699	Standard
	Be	9	20.0	25.0	**0.0316**	0.005	17.2	ug/L	8	Standard
	Al	27	2221.8	11.1	**0.0027**	0.002	62.7	ug/L	1493	Standard
	Sc	45	58138.1	3.2				ug/L	61029	Standard
	Ti	47	47.0	20.3	**0.0161**	0.026	158.4	ug/L	35	Standard
	V	51	1216.9	5.3	**-0.0330**	0.008	24.4	ug/L	1435	Standard
	Cr	52	6478.4	1.5	**-0.1416**	0.009	6.0	ug/L	7635	Standard
	Cr	53	358.3	10.7	**0.0617**	0.046	74.0	ug/L	363	Standard
	Mn	55	785.0	2.1	**-0.0343**	0.003	7.5	ug/L	747	Standard
	Co	59	292.7	8.2	**0.0075**	0.004	50.2	ug/L	234	Standard
	Ni	60	209.7	4.3	**-0.0109**	0.006	51.3	ug/L	215	Standard
	Cu	65	121.0	5.8	**-0.0381**	0.003	8.4	ug/L	104	Standard
	Zn	66	398.3	6.4	**-0.0865**	0.023	26.3	ug/L	363	Standard
>	**Ge**	72	490643.1	1.0				ug/L	495142	Standard
	As	75	-2.3	851.2	**0.0577**	0.017	30.2	ug/L	-50	Standard
	Se	82	15.9	39.8	**0.0087**	0.059	680.3	ug/L	14	Standard
	Se-1	77	51.3	11.1	**0.0869**	0.077	88.1	ug/L	42	Standard
>	**Ga**	71	46.7	16.4				mg/L	42	Standard
	Rb	85	33.3	45.8				ug/L	17	Standard
	Y	89	424498.7	1.6				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	405.0	23.6	**0.1487**	0.033	22.3	ug/L	11	Standard
	Ag	107	128.3	5.2	**0.0071**	0.001	8.6	ug/L	68	Standard
	Cd	111	15.0	38.7	**-0.0006**	0.003	474.1	mg/L	2	Standard
	Cd	114	47.2	8.1	**0.0073**	0.001	11.6	ug/L	12	Standard
>	**In**	115	576512.9	2.1				ug/L	587441	Standard
	Sn	118	851.7	19.9	**0.0543**	0.032	58.8	ug/L	497	Standard
	Sb	123	1550.0	28.1	**0.3352**	0.100	29.7	ug/L	78	Standard
	Ba	135	36.7	23.2	**0.0022**	0.004	189.8	ug/L	24	Standard
	Ce	140	21.7	93.3				ug/L	25	Standard
>	**Tb**	159	924023.3	3.2				ug/L	939701	Standard
	Ho	165	10.0	100.0				ug/L	10	Standard
	Tl	203	60.7	18.1	**0.0057**	0.001	22.9	ug/L	4	Standard
	Tl	205	61.7	61.4	**0.0089**	0.005	59.3	ug/L	2	Standard
	Pb	206	271.3	7.0	**-0.0198**	0.004	19.8	ug/L	240	Standard
	Pb	207	237.3	7.1	**-0.0194**	0.004	21.0	ug/L	198	Standard
	Pb	208	1028.3	6.0	**-0.0187**	0.003	15.4	ug/L	883	Standard
	U	238	47.7	28.5	**0.0026**	0.002	78.0	ug/L	3	Standard
>	**Bi**	209	518292.4	0.9				ug/L	523984	Standard

Sample ID: QC Std 2
Report Date/Time: Wednesday, June 29, 2016 11:44:03
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	35.0	42.9	**-0.0179**	0.019	107.8	mg/L	37	Standard
K	39	36.7	39.4	**-0.0284**	0.035	124.9	mg/L	38	Standard
Ca	43	25.0	40.0	**-0.8569**	0.683	79.6	mg/L	10	Standard
Fe	54	125.6	8.4	**-0.0262**	0.013	49.1	mg/L	179	Standard
Fe	57	296.7	4.9	**-0.0150**	0.061	408.2	mg/L	335	Standard
> **Sc-1**	45	58138.1	3.2				mg/L	61029	Standard
Cl	35	20966.5	2.9				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	1540.1	1.7				ug/L	1547	Standard
P	31	17027.9	2.4				ug/L	15286	Standard
S	34	1976.8	0.6				ug/L	1572	Standard
Sr	88	85.0	5.9				ug/L	73	Standard
C	12	30.0	33.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	6.3	86.9				mg/L	6	Standard
Ho-1	165	10.0	100.0				mg/L	10	Standard
Er	166	6.7	173.2				mg/L	13	Standard
I	127	1458.4	25.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.091	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 2
Report Date/Time: Wednesday, June 29, 2016 11:44:03
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.140
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.914
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 2	Sb	123	
QC Std 2	Ca	43	

Sample ID: QC Std 2
Report Date/Time: Wednesday, June 29, 2016 11:44:03
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 3
Sample Date/Time: Wednesday, June 29, 2016 11:46:53
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	66549.2	2.6				ug/L	74699	Standard
	Be	9	181.7	3.2	**0.2393**	0.009	3.6	ug/L	8	Standard
	Al	27	1430.1	59.0	**-0.0024**	0.007	267.2	ug/L	1493	Standard
	Sc	45	54693.6	2.6				ug/L	61029	Standard
	Ti	47	40.3	11.5	**0.0025**	0.013	506.9	ug/L	35	Standard
	V	51	4161.2	3.6	**0.3819**	0.016	4.1	ug/L	1435	Standard
	Cr	52	12463.1	1.9	**0.7390**	0.018	2.5	ug/L	7635	Standard
	Cr	53	996.7	1.0	**0.8006**	0.013	1.6	ug/L	363	Standard
	Mn	55	4091.6	2.3	**0.4679**	0.008	1.8	ug/L	747	Standard
	Co	59	2840.6	1.4	**0.4137**	0.010	2.4	ug/L	234	Standard
	Ni	60	3014.0	1.2	**1.5638**	0.035	2.2	ug/L	215	Standard
	Cu	65	1546.7	1.4	**0.7676**	0.020	2.7	ug/L	104	Standard
	Zn	66	7394.5	0.3	**6.6420**	0.044	0.7	ug/L	363	Standard
>	**Ge**	72	472336.8	0.9				ug/L	495142	Standard
	As	75	394.7	6.3	**0.4290**	0.024	5.7	ug/L	-50	Standard
	Se	82	54.4	27.0	**0.3818**	0.143	37.5	ug/L	14	Standard
	Se-1	77	75.3	14.2	**0.4599**	0.162	35.3	ug/L	42	Standard
>	**Ga**	71	58.3	44.0				mg/L	42	Standard
	Rb	85	8.3	34.6				ug/L	17	Standard
	Y	89	402582.8	1.9				ug/L	417098	Standard
>	**Rh**	103	5.0	100.0				ug/L	13	Standard
	Mo	98	114.4	34.0	**0.0400**	0.016	39.3	ug/L	11	Standard
	Ag	107	2464.9	1.7	**0.3900**	0.008	2.1	ug/L	68	Standard
	Cd	111	479.8	5.6	**0.2365**	0.013	5.7	mg/L	2	Standard
	Cd	114	1217.0	2.5	**0.2507**	0.007	2.6	ug/L	12	Standard
>	**In**	115	551471.6	0.4				ug/L	587441	Standard
	Sn	118	548.3	23.3	**0.0053**	0.024	443.1	ug/L	497	Standard
	Sb	123	2017.6	6.1	**0.4562**	0.029	6.4	ug/L	78	Standard
	Ba	135	1439.7	3.7	**0.7333**	0.029	4.0	ug/L	24	Standard
	Ce	140	11.7	49.5				ug/L	25	Standard
>	**Tb**	159	887425.9	1.8				ug/L	939701	Standard
	Ho	165	10.0	50.0				ug/L	10	Standard
	Tl	203	701.7	6.6	**0.0871**	0.005	5.3	ug/L	4	Standard
	Tl	205	588.3	14.1	**0.0861**	0.013	15.2	ug/L	2	Standard
	Pb	206	1333.4	2.9	**0.1894**	0.006	3.1	ug/L	240	Standard
	Pb	207	1083.0	1.4	**0.1683**	0.005	3.1	ug/L	198	Standard
	Pb	208	4575.3	2.2	**0.1806**	0.004	2.0	ug/L	883	Standard
	U	238	2677.9	3.2	**0.4017**	0.012	2.9	ug/L	3	Standard
>	**Bi**	209	501924.0	1.5				ug/L	523984	Standard

Sample ID: QC Std 3
Report Date/Time: Wednesday, June 29, 2016 11:49:09
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**21.9523**	38.014	173.2	mg/L	0	Standard
Mg	24	43.3	29.0	**-0.0023**	0.020	841.8	mg/L	37	Standard
K	39	33.3	37.7	**-0.0317**	0.032	101.1	mg/L	38	Standard
Ca	43	46.7	12.4	**0.7738**	0.482	62.2	mg/L	10	Standard
Fe	54	100.7	18.6	**-0.0432**	0.016	36.6	mg/L	179	Standard
Fe	57	325.0	8.1	**0.1578**	0.085	53.5	mg/L	335	Standard
> **Sc-1**	45	54693.6	2.6				mg/L	61029	Standard
Cl	35	21199.5	1.6				ug/L	20195	Standard
Kr	83	3.3	17.3				ug/L	4	Standard
Br	81	1403.4	5.3				ug/L	1547	Standard
P	31	11833.0	1.6				ug/L	15286	Standard
S	34	2000.1	8.0				ug/L	1572	Standard
Sr	88	93.3	20.3				ug/L	73	Standard
C	12	56.7	10.2				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	13.0	89.8				mg/L	6	Standard
Ho-1	165	10.0	50.0				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1430.1	34.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	119.656		
Al	27	-0.244		
Sc	45			
Ti	47			
V	51	95.473		
Cr	52	92.381		
Cr	53			
Mn	55	93.583		
Co	59	103.429		
Ni	60	97.737		
Cu	65	95.951		
Zn	66	106.273		
> Ge	72		95.394	
As	75	107.259		
Se	82	95.456		
Se-1	77			
> Ga	71			

Sample ID: QC Std 3
Report Date/Time: Wednesday, June 29, 2016 11:49:09
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	97.496		
Cd	111	98.556		
Cd	114			
> In	115			93.877
Sn	118			
Sb	123	114.045		
Ba	135	97.779		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	108.827		
Tl	205			
Pb	206			
Pb	207			
Pb	208	90.316		
U	238	100.429		
> Bi	209			95.790
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 3	Al	27	

Sample ID: QC Std 3
Report Date/Time: Wednesday, June 29, 2016 11:49:09
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Wednesday, June 29, 2016 11:50:53
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	71628.9	1.8				ug/L	74699	Standard
	Be	9	13.3	43.3	**0.0235**	0.007	30.3	ug/L	8	Standard
	Al	27	7079214.4	1.4	**50.2405**	1.571	3.1	ug/L	1493	Standard
	Sc	45	57092.4	1.6				ug/L	61029	Standard
	Ti	47	41841.7	1.2	**112.2257**	1.296	1.2	ug/L	35	Standard
	V	51	1975.7	11.7	**0.0706**	0.031	44.3	ug/L	1435	Standard
	Cr	52	7951.4	1.4	**0.0680**	0.015	21.7	ug/L	7635	Standard
	Cr	53	3292.0	3.5	**3.2946**	0.126	3.8	ug/L	363	Standard
	Mn	55	1443.4	4.1	**0.0629**	0.009	13.6	ug/L	747	Standard
	Co	59	431.3	5.1	**0.0292**	0.003	11.6	ug/L	234	Standard
	Ni	60	837.4	1.6	**0.3312**	0.007	2.2	ug/L	215	Standard
	Cu	65	638.0	4.2	**0.2451**	0.015	6.0	ug/L	104	Standard
	Zn	66	1716.8	2.4	**1.1448**	0.039	3.4	ug/L	363	Standard
>	Ge	72	486594.7	0.0				ug/L	495142	Standard
	As	75	-47.8	111.3	**0.0161**	0.048	300.1	ug/L	-50	Standard
	Se	82	12.5	17.3	**-0.0221**	0.020	90.2	ug/L	14	Standard
	Se-1	77	285.3	3.8	**3.3530**	0.150	4.5	ug/L	42	Standard
>	Ga	71	146.7	5.2				mg/L	42	Standard
	Rb	85	1056.7	1.8				ug/L	17	Standard
	Y	89	409909.2	0.2				ug/L	417098	Standard
>	Rh	103	11.7	99.0				ug/L	13	Standard
	Mo	98	258876.2	1.4	**101.0110**	2.466	2.4	ug/L	11	Standard
	Ag	107	250.7	8.2	**0.0271**	0.004	13.5	ug/L	68	Standard
	Cd	111	-191.4	6.3	**-0.1031**	0.005	5.1	mg/L	2	Standard
	Cd	114	476.2	9.0	**0.0945**	0.009	9.4	ug/L	12	Standard
>	In	115	564912.9	1.1				ug/L	587441	Standard
	Sn	118	765.0	2.0	**0.0415**	0.002	5.8	ug/L	497	Standard
	Sb	123	314.8	31.7	**0.0658**	0.022	32.8	ug/L	78	Standard
	Ba	135	628.0	0.0	**0.3031**	0.004	1.2	ug/L	24	Standard
	Ce	140	243.3	10.5				ug/L	25	Standard
>	Tb	159	934364.8	1.3				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	234.7	3.6	**0.0271**	0.001	3.8	ug/L	4	Standard
	Tl	205	175.0	22.3	**0.0250**	0.006	22.8	ug/L	2	Standard
	Pb	206	514.7	3.8	**0.0265**	0.004	16.6	ug/L	240	Standard
	Pb	207	427.7	5.6	**0.0214**	0.005	22.5	ug/L	198	Standard
	Pb	208	1691.4	2.8	**0.0174**	0.002	11.4	ug/L	883	Standard
	U	238	20.7	10.1	**-0.0014**	0.000	22.7	ug/L	3	Standard
>	Bi	209	516629.8	0.9				ug/L	523984	Standard

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 11:53:10
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**102.8074**	36.030	35.0	mg/L	0	Standard
Mg	24	9604.7	1.9	**13.1909**	0.366	2.8	mg/L	37	Standard
K	39	2033.5	6.6	**4.6893**	0.286	6.1	mg/L	38	Standard
Ca	43	250.0	4.0	**14.3615**	0.772	5.4	mg/L	10	Standard
Fe	54	14441.1	1.7	**12.9538**	0.416	3.2	mg/L	179	Standard
Fe	57	3845.5	3.6	**12.9047**	0.714	5.5	mg/L	335	Standard
> **Sc-1**	45	57092.4	1.6				mg/L	61029	Standard
Cl	35	20428.4	2.1				ug/L	20195	Standard
Kr	83	3.0	57.7				ug/L	4	Standard
Br	81	1493.4	8.7				ug/L	1547	Standard
P	31	15903.3	2.0				ug/L	15286	Standard
S	34	1818.4	8.5				ug/L	1572	Standard
Sr	88	70.0	35.7				ug/L	73	Standard
C	12	136.7	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	46.7	86.6				mg/L	3	Standard
Dy	164	12.2	93.4				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	23.3	24.7				mg/L	13	Standard
I	127	2226.8	19.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27	1.005		
Sc	45			
Ti	47	112.226		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.274	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 11:53:10
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	101.011		
Ag	107			
Cd	111			
Cd	114			
> In	115		96.165	
Sn	118			
Sb	123			
Ba	135			
Ce	140			
> Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
> Bi	209		98.596	
Na	23	822.459		
Mg	24	263.819		
K	39	93.787		
Ca	43	95.743		
Fe	54	103.630		
Fe	57	103.238		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Na	23	
QC Std 4	Mg	24	

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 11:53:10
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Wednesday, June 29, 2016 11:54:47
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	72764.4	2.4				ug/L	74699	Standard
	Be	9	86093.9	0.6	**100.4334**	2.352	2.3	ug/L	8	Standard
	Al	27	7577089.0	1.5	**52.9410**	1.751	3.3	ug/L	1493	Standard
	Sc	45	54926.1	3.1				ug/L	61029	Standard
	Ti	47	43156.4	1.1	**115.6496**	0.651	0.6	ug/L	35	Standard
	V	51	767037.5	0.7	**103.0563**	0.476	0.5	ug/L	1435	Standard
	Cr	52	743370.8	0.7	**100.9527**	0.091	0.1	ug/L	7635	Standard
	Cr	53	95736.5	0.4	**104.9834**	0.488	0.5	ug/L	363	Standard
	Mn	55	719673.0	1.2	**104.9487**	1.961	1.9	ug/L	747	Standard
	Co	59	684381.8	1.6	**105.3392**	2.464	2.3	ug/L	234	Standard
	Ni	60	184775.9	0.8	**100.2112**	0.878	0.9	ug/L	215	Standard
	Cu	65	183945.6	0.7	**100.3790**	1.285	1.3	ug/L	104	Standard
	Zn	66	111716.4	0.5	**103.5231**	0.465	0.4	ug/L	363	Standard
>	Ge	72	487033.1	0.8				ug/L	495142	Standard
	As	75	115123.8	0.6	**104.5507**	0.234	0.2	ug/L	-50	Standard
	Se	82	11162.3	1.3	**103.2681**	1.937	1.9	ug/L	14	Standard
	Se-1	77	7790.3	1.4	**107.8133**	0.940	0.9	ug/L	42	Standard
>	Ga	71	155.0	24.4				mg/L	42	Standard
	Rb	85	853.4	12.0				ug/L	17	Standard
	Y	89	414937.1	1.1				ug/L	417098	Standard
>	Rh	103	48.3	21.5				ug/L	13	Standard
	Mo	98	271334.5	0.2	**106.0489**	0.711	0.7	ug/L	11	Standard
	Ag	107	596492.9	1.3	**95.3413**	1.752	1.8	ug/L	68	Standard
	Cd	111	202245.3	1.2	**100.7713**	1.934	1.9	mg/L	2	Standard
	Cd	114	489634.1	0.4	**99.4602**	0.649	0.7	ug/L	12	Standard
>	In	115	563893.1	0.8				ug/L	587441	Standard
	Sn	118	898.4	12.5	**0.0654**	0.019	28.7	ug/L	497	Standard
	Sb	123	448934.9	1.4	**100.1777**	2.039	2.0	ug/L	78	Standard
	Ba	135	193320.5	0.8	**98.3885**	1.494	1.5	ug/L	24	Standard
	Ce	140	68.3	16.9				ug/L	25	Standard
>	Tb	159	922182.5	1.6				ug/L	939701	Standard
	Ho	165	15.0	57.7				ug/L	10	Standard
	Tl	203	808287.0	0.7	**100.1763**	1.478	1.5	ug/L	4	Standard
	Tl	205	710155.2	1.4	**101.4182**	2.560	2.5	ug/L	2	Standard
	Pb	206	542765.1	0.5	**103.7300**	0.986	1.0	ug/L	240	Standard
	Pb	207	466015.8	1.0	**100.2432**	1.559	1.6	ug/L	198	Standard
	Pb	208	1879754.6	0.3	**102.3429**	1.262	1.2	ug/L	883	Standard
	U	238	673426.3	1.1	**99.9548**	2.202	2.2	ug/L	3	Standard
>	Bi	209	512933.6	1.2				ug/L	523984	Standard

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 11:57:04
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	10138.4	1.8	**14.4874**	0.596	4.1	mg/L	37	Standard
K	39	2125.1	3.3	**5.1060**	0.147	2.9	mg/L	38	Standard
Ca	43	265.0	8.2	**16.1201**	2.142	13.3	mg/L	10	Standard
Fe	54	14876.0	0.8	**13.8828**	0.353	2.5	mg/L	179	Standard
Fe	57	3958.9	4.0	**13.8986**	1.069	7.7	mg/L	335	Standard
Sc-1	45	54926.1	3.1				mg/L	61029	Standard
Cl	35	20351.0	1.2				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	1600.1	9.2				ug/L	1547	Standard
P	31	14895.7	0.4				ug/L	15286	Standard
S	34	1618.4	8.3				ug/L	1572	Standard
Sr	88	85.0	10.2				ug/L	73	Standard
C	12	116.7	71.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	16.2	38.3				mg/L	6	Standard
Ho-1	165	15.0	57.7				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	1496.7	25.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	100.433		
Al	27	1.059		
Sc	45			
Ti	47	115.650		
V	51	103.056		
Cr	52	100.953		
Cr	53			
Mn	55	104.949		
Co	59	105.339		
Ni	60	100.211		
Cu	65	100.379		
Zn	66	103.523		
Ge	72		98.362	
As	75	104.551		
Se	82	103.268		
Se-1	77			
Ga	71			

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 11:57:04
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	106.049		
Ag	107	95.341		
Cd	111	100.771		
Cd	114			
> In	115		95.991	
Sn	118			
Sb	123	100.178		
Ba	135	98.389		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	100.176		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.343		
U	238	99.955		
> Bi	209		97.891	
Na	23	0.040		
Mg	24	289.748		
K	39	102.121		
Ca	43	107.468		
Fe	54	111.063		
Fe	57	111.189		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Na	23	
QC Std 5	Mg	24	

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 11:57:04
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 12:00:10
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	74127.8	4.1				ug/L	74699	Standard
	Be	9	44057.3	2.0	**50.4723**	1.532	3.0	ug/L	8	Standard
	Al	27	7182862.5	0.8	**49.2846**	1.691	3.4	ug/L	1493	Standard
	Sc	45	54972.9	1.9				ug/L	61029	Standard
	Ti	47	37035.5	1.0	**100.2235**	3.375	3.4	ug/L	35	Standard
	V	51	380987.7	1.2	**51.6037**	1.920	3.7	ug/L	1435	Standard
	Cr	52	371095.9	0.5	**50.3905**	1.647	3.3	ug/L	7635	Standard
	Cr	53	45648.7	0.9	**50.3715**	0.927	1.8	ug/L	363	Standard
	Mn	55	362758.6	1.1	**53.3556**	2.033	3.8	ug/L	747	Standard
	Co	59	346170.1	1.6	**53.7955**	2.203	4.1	ug/L	234	Standard
	Ni	60	91822.1	1.5	**50.2407**	2.119	4.2	ug/L	215	Standard
	Cu	65	92043.5	2.0	**50.6873**	2.359	4.7	ug/L	104	Standard
	Zn	66	54708.9	1.1	**50.9762**	1.937	3.8	ug/L	363	Standard
>	Ge	72	482505.9	2.7				ug/L	495142	Standard
	As	75	56019.2	1.2	**51.4117**	1.775	3.5	ug/L	-50	Standard
	Se	82	5615.5	1.4	**52.4035**	2.100	4.0	ug/L	14	Standard
	Se-1	77	3787.5	1.5	**52.6293**	2.112	4.0	ug/L	42	Standard
>	Ga	71	63.3	9.1				mg/L	42	Standard
	Rb	85	771.7	5.9				ug/L	17	Standard
	Y	89	419939.4	2.7				ug/L	417098	Standard
>	Rh	103	31.7	9.1				ug/L	13	Standard
	Mo	98	262674.3	2.0	**102.5093**	6.034	5.9	ug/L	11	Standard
	Ag	107	318557.4	0.8	**50.8058**	2.007	4.0	ug/L	68	Standard
	Cd	111	103808.6	0.6	**51.6156**	2.157	4.2	mg/L	2	Standard
	Cd	114	253134.5	2.1	**51.2889**	1.034	2.0	ug/L	12	Standard
>	In	115	565595.5	4.0				ug/L	587441	Standard
	Sn	118	290084.1	1.0	**51.5381**	2.174	4.2	ug/L	497	Standard
	Sb	123	228198.9	1.3	**50.8307**	2.651	5.2	ug/L	78	Standard
	Ba	135	99366.1	0.6	**50.4579**	1.943	3.9	ug/L	24	Standard
	Ce	140	78.3	13.3				ug/L	25	Standard
>	Tb	159	913248.6	2.6				ug/L	939701	Standard
	Ho	165	3.3	86.6				ug/L	10	Standard
	Tl	203	413066.5	1.0	**51.9084**	1.016	2.0	ug/L	4	Standard
	Tl	205	363184.8	0.9	**52.5982**	1.678	3.2	ug/L	2	Standard
	Pb	206	268800.2	0.8	**52.0649**	1.553	3.0	ug/L	240	Standard
	Pb	207	239468.7	1.2	**52.2107**	1.790	3.4	ug/L	198	Standard
	Pb	208	958802.6	0.9	**52.9068**	1.662	3.1	ug/L	883	Standard
	U	238	349772.5	0.2	**52.6413**	1.317	2.5	ug/L	3	Standard
>	Bi	209	505954.2	2.3				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:02:27
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	3805.5	5.3	**5.3881**	0.260	4.8	mg/L	37	Standard
K	39	2106.8	1.9	**5.0574**	0.179	3.5	mg/L	38	Standard
Ca	43	101.7	5.7	**4.6110**	0.501	10.9	mg/L	10	Standard
Fe	54	5879.5	1.7	**5.3975**	0.172	3.2	mg/L	179	Standard
Fe	57	1785.1	7.3	**5.6694**	0.625	11.0	mg/L	335	Standard
> **Sc-1**	45	54972.9	1.9				mg/L	61029	Standard
Cl	35	21821.7	0.9				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	1470.1	4.5				ug/L	1547	Standard
P	31	16677.5	6.0				ug/L	15286	Standard
S	34	1955.1	4.5				ug/L	1572	Standard
Sr	88	73.3	14.2				ug/L	73	Standard
C	12	80.0	66.1				mg/L	63	Standard
N	14	6.7	86.6				mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.9	32.3				mg/L	6	Standard
Ho-1	165	3.3	86.6				mg/L	10	Standard
Er	166	16.7	124.9				mg/L	13	Standard
I	127	1668.4	28.6				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	100.945		
	Al	27	98.569		
	Sc	45			
	Ti	47	100.224		
	V	51	103.207		
	Cr	52	100.781		
	Cr	53			
	Mn	55	106.711		
	Co	59	107.591		
	Ni	60	100.481		
	Cu	65	101.375		
	Zn	66	101.952		
>	Ge	72		97.448	
	As	75	102.823		
	Se	82	104.807		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:02:27
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.509		
Ag	107	101.612		
Cd	111	103.231		
Cd	114			
In	115			96.281
Sn	118	103.076		
Sb	123	101.661		
Ba	135	100.916		
Ce	140			
Tb	159			
Ho	165			
Tl	203	103.817		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.814		
U	238	105.283		
Bi	209			96.559
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:02:27
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 12:07:34
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74348.6	1.0				ug/L	74699	Standard
	Be	9	21.7	26.6	**0.0324**	0.007	20.7	ug/L	8	Standard
	Al	27	1840.1	21.7	**-0.0008**	0.003	358.1	ug/L	1493	Standard
	Sc	45	56396.5	2.8				ug/L	61029	Standard
	Ti	47	36.7	15.0	**-0.0098**	0.014	147.0	ug/L	35	Standard
	V	51	1326.6	5.0	**-0.0153**	0.008	51.0	ug/L	1435	Standard
	Cr	52	6702.5	2.0	**-0.0955**	0.015	16.1	ug/L	7635	Standard
	Cr	53	393.3	15.3	**0.1070**	0.067	63.0	ug/L	363	Standard
	Mn	55	742.7	5.5	**-0.0386**	0.005	13.8	ug/L	747	Standard
	Co	59	250.3	1.5	**0.0016**	0.000	18.8	ug/L	234	Standard
	Ni	60	195.0	9.2	**-0.0170**	0.010	58.7	ug/L	215	Standard
	Cu	65	108.7	1.4	**-0.0438**	0.000	1.1	ug/L	104	Standard
	Zn	66	377.3	7.3	**-0.1000**	0.024	24.2	ug/L	363	Standard
>	**Ge**	72	482419.6	0.7				ug/L	495142	Standard
	As	75	-16.3	122.1	**0.0447**	0.018	40.8	ug/L	-50	Standard
	Se	82	12.2	10.1	**-0.0240**	0.012	49.4	ug/L	14	Standard
	Se-1	77	52.7	12.2	**0.1177**	0.089	75.6	ug/L	42	Standard
>	**Ga**	71	61.7	33.8				mg/L	42	Standard
	Rb	85	30.0	50.0				ug/L	17	Standard
	Y	89	414790.7	2.9				ug/L	417098	Standard
>	**Rh**	103	1.7	173.2				ug/L	13	Standard
	Mo	98	135.9	9.5	**0.0469**	0.005	10.2	ug/L	11	Standard
	Ag	107	96.3	7.3	**0.0024**	0.001	45.2	ug/L	68	Standard
	Cd	111	7.4	43.0	**-0.0042**	0.002	37.1	mg/L	2	Standard
	Cd	114	24.8	51.6	**0.0029**	0.003	88.9	ug/L	12	Standard
>	**In**	115	568967.0	0.5				ug/L	587441	Standard
	Sn	118	628.3	16.1	**0.0164**	0.018	110.8	ug/L	497	Standard
	Sb	123	617.2	40.6	**0.1324**	0.056	42.3	ug/L	78	Standard
	Ba	135	30.7	16.1	**-0.0005**	0.003	469.7	ug/L	24	Standard
	Ce	140	23.3	68.9				ug/L	25	Standard
>	**Tb**	159	918183.5	1.2				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	30.3	14.9	**0.0020**	0.001	31.1	ug/L	4	Standard
	Tl	205	23.3	24.7	**0.0035**	0.001	22.2	ug/L	2	Standard
	Pb	206	259.3	0.6	**-0.0221**	0.001	4.2	ug/L	240	Standard
	Pb	207	198.7	1.6	**-0.0277**	0.001	4.0	ug/L	198	Standard
	Pb	208	894.0	2.8	**-0.0259**	0.002	7.7	ug/L	883	Standard
	U	238	15.0	35.3	**-0.0022**	0.001	35.0	ug/L	3	Standard
>	**Bi**	209	518580.7	1.8				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:09:51
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**20.6253**	35.715	173.2	mg/L	0	Standard
Mg	24	31.7	32.9	**-0.0207**	0.015	73.7	mg/L	37	Standard
K	39	51.7	24.4	**0.0090**	0.027	299.1	mg/L	38	Standard
Ca	43	33.3	31.2	**-0.2319**	0.781	336.6	mg/L	10	Standard
Fe	54	140.8	7.6	**-0.0093**	0.006	67.4	mg/L	179	Standard
Fe	57	256.7	18.9	**-0.1295**	0.180	138.6	mg/L	335	Standard
> **Sc-1**	45	56396.5	2.8				mg/L	61029	Standard
Cl	35	21372.4	1.3				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	1540.1	8.8				ug/L	1547	Standard
P	31	15931.7	0.9				ug/L	15286	Standard
S	34	1863.4	1.4				ug/L	1572	Standard
Sr	88	58.3	17.8				ug/L	73	Standard
C	12	56.7	27.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	6.2	186.6				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	10.0					mg/L	13	Standard
I	127	1518.4	26.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.431	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:09:51
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.855
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.969
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:09:51
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066106
Sample Date/Time: Wednesday, June 29, 2016 12:12:15
Number of Replicates: 3
Autosampler Position: 301
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78301.9	3.0				ug/L	74699	Standard
	Be	9	591.7	15.8	**0.6482**	0.094	14.5	ug/L	8	Standard
	Al	27	14033022.6	1.2	**91.1117**	1.566	1.7	ug/L	1493	Standard
	Sc	45	64031.7	1.2				ug/L	61029	Standard
	Ti	47	6936.9	1.2	**18.7002**	0.397	2.1	ug/L	35	Standard
	V	51	136379.5	0.6	**18.3618**	0.342	1.9	ug/L	1435	Standard
	Cr	52	135090.6	0.2	**17.7119**	0.504	2.8	ug/L	7635	Standard
	Cr	53	16100.2	3.4	**17.5840**	1.043	5.9	ug/L	363	Standard
	Mn	55	2727656.7	0.9	**402.5207**	10.967	2.7	ug/L	747	Standard
	Co	59	46530.0	0.2	**7.2050**	0.195	2.7	ug/L	234	Standard
	Ni	60	31243.2	1.0	**17.0295**	0.585	3.4	ug/L	215	Standard
	Cu	65	40278.5	0.5	**22.1396**	0.604	2.7	ug/L	104	Standard
	Zn	66	62615.1	0.3	**58.4626**	1.558	2.7	ug/L	363	Standard
>	**Ge**	72	481946.2	2.4				ug/L	495142	Standard
	As	75	4644.7	1.4	**4.3222**	0.154	3.6	ug/L	-50	Standard
	Se	82	54.0	14.6	**0.3688**	0.086	23.4	ug/L	14	Standard
	Se-1	77	96.0	5.2	**0.7275**	0.040	5.5	ug/L	42	Standard
>	**Ga**	71	12810.4	4.6				mg/L	42	Standard
	Rb	85	116256.0	0.7				ug/L	17	Standard
	Y	89	585719.0	1.0				ug/L	417098	Standard
>	**Rh**	103	45.0	33.3				ug/L	13	Standard
	Mo	98	935.2	3.6	**0.3550**	0.010	2.8	ug/L	11	Standard
	Ag	107	931.4	1.2	**0.1340**	0.001	0.5	ug/L	68	Standard
	Cd	111	782.7	0.6	**0.3771**	0.007	1.8	mg/L	2	Standard
	Cd	114	1959.7	7.4	**0.3906**	0.025	6.5	ug/L	12	Standard
>	**In**	115	571359.2	1.2				ug/L	587441	Standard
	Sn	118	1011.7	7.7	**0.0833**	0.012	14.7	ug/L	497	Standard
	Sb	123	336.5	28.9	**0.0698**	0.021	29.4	ug/L	78	Standard
	Ba	135	394959.7	1.6	**198.3814**	1.476	0.7	ug/L	24	Standard
	Ce	140	893729.8	1.7				ug/L	25	Standard
>	**Tb**	159	948738.0	1.3				ug/L	939701	Standard
	Ho	165	15724.8	2.6				ug/L	10	Standard
	Tl	203	658.0	2.7	**0.0791**	0.002	2.2	ug/L	4	Standard
	Tl	205	583.3	11.6	**0.0828**	0.010	12.2	ug/L	2	Standard
	Pb	206	370590.1	1.2	**70.2116**	0.816	1.2	ug/L	240	Standard
	Pb	207	304396.4	0.8	**64.9093**	0.988	1.5	ug/L	198	Standard
	Pb	208	1254498.7	1.7	**67.7133**	1.716	2.5	ug/L	883	Standard
	U	238	3382.4	2.2	**0.4935**	0.016	3.3	ug/L	3	Standard
>	**Bi**	209	517229.2	1.1				ug/L	523984	Standard

Sample ID: L1606066106
Report Date/Time: Wednesday, June 29, 2016 12:14:32
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**36.6019**	63.388	173.2	mg/L	0	Standard
Mg	24	128.3	15.7	**0.0927**	0.024	25.5	mg/L	37	Standard
K	39	566.7	14.3	**1.0807**	0.180	16.7	mg/L	38	Standard
Ca	43	418.3	12.7	**22.6893**	3.491	15.4	mg/L	10	Standard
Fe	54	12793.3	3.0	**10.1983**	0.198	1.9	mg/L	179	Standard
Fe	57	3678.8	1.9	**10.8461**	0.363	3.4	mg/L	335	Standard
Sc-1	45	64031.7	1.2				mg/L	61029	Standard
Cl	35	22050.7	1.1				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2173.5	11.7				ug/L	1547	Standard
P	31	16252.1	5.4				ug/L	15286	Standard
S	34	2065.1	4.4				ug/L	1572	Standard
Sr	88	140.0	9.4				ug/L	73	Standard
C	12	143.3	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	203.3	10.2				mg/L	3	Standard
Dy	164	25060.4	3.8				mg/L	6	Standard
Ho-1	165	15724.8	2.6				mg/L	10	Standard
Er	166	13529.4	4.1				mg/L	13	Standard
I	127	8297.3	2.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		104.824	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		97.335	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606066106
Report Date/Time: Wednesday, June 29, 2016 12:14:32
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.262
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	98.711
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606066106
Report Date/Time: Wednesday, June 29, 2016 12:14:32
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066107
Sample Date/Time: Wednesday, June 29, 2016 12:15:26
Number of Replicates: 3
Autosampler Position: 302
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79855.0	3.3				ug/L	74699	Standard
	Be	9	765.0	6.8	**0.8219**	0.077	9.4	ug/L	8	Standard
	Al	27	26230756.6	0.7	**167.0452**	4.333	2.6	ug/L	1493	Standard
	Sc	45	64585.8	4.1				ug/L	61029	Standard
	Ti	47	7966.1	1.8	**21.8574**	0.352	1.6	ug/L	35	Standard
	V	51	155637.7	1.4	**21.3440**	0.459	2.1	ug/L	1435	Standard
	Cr	52	137713.7	0.8	**18.3989**	0.313	1.7	ug/L	7635	Standard
	Cr	53	16806.0	3.9	**18.6780**	0.841	4.5	ug/L	363	Standard
	Mn	55	4726467.2	1.1	**709.3402**	2.876	0.4	ug/L	747	Standard
	Co	59	64550.2	0.1	**10.1794**	0.081	0.8	ug/L	234	Standard
	Ni	60	42709.8	0.1	**23.7183**	0.226	1.0	ug/L	215	Standard
	Cu	65	23683.8	1.0	**13.1970**	0.251	1.9	ug/L	104	Standard
	Zn	66	44795.8	1.0	**42.4101**	0.817	1.9	ug/L	363	Standard
>	**Ge**	72	473762.9	0.9				ug/L	495142	Standard
	As	75	7180.0	1.6	**6.7588**	0.066	1.0	ug/L	-50	Standard
	Se	82	142.2	5.2	**1.2157**	0.067	5.5	ug/L	14	Standard
	Se-1	77	110.3	7.6	**0.9569**	0.126	13.2	ug/L	42	Standard
>	**Ga**	71	17858.9	0.4				mg/L	42	Standard
	Rb	85	152815.2	1.7				ug/L	17	Standard
	Y	89	630725.5	1.5				ug/L	417098	Standard
>	**Rh**	103	41.7	25.0				ug/L	13	Standard
	Mo	98	1176.0	1.0	**0.4476**	0.010	2.2	ug/L	11	Standard
	Ag	107	471.0	1.7	**0.0613**	0.002	3.5	ug/L	68	Standard
	Cd	111	231.8	9.8	**0.1061**	0.013	12.1	mg/L	2	Standard
	Cd	114	548.2	12.6	**0.1078**	0.015	13.9	ug/L	12	Standard
>	**In**	115	571871.5	1.5				ug/L	587441	Standard
	Sn	118	1063.4	7.5	**0.0923**	0.015	16.1	ug/L	497	Standard
	Sb	123	198.5	41.2	**0.0396**	0.018	46.2	ug/L	78	Standard
	Ba	135	410562.2	0.2	**206.0643**	2.723	1.3	ug/L	24	Standard
	Ce	140	1020863.3	0.8				ug/L	25	Standard
>	**Tb**	159	935971.5	1.4				ug/L	939701	Standard
	Ho	165	18616.4	1.9				ug/L	10	Standard
	Tl	203	969.7	3.1	**0.1217**	0.006	4.6	ug/L	4	Standard
	Tl	205	781.7	5.1	**0.1148**	0.005	4.1	ug/L	2	Standard
	Pb	206	74583.8	0.8	**14.5751**	0.129	0.9	ug/L	240	Standard
	Pb	207	60293.4	1.5	**13.2552**	0.050	0.4	ug/L	198	Standard
	Pb	208	253434.0	0.7	**14.1059**	0.309	2.2	ug/L	883	Standard
	U	238	4763.7	0.8	**0.7216**	0.012	1.7	ug/L	3	Standard
>	**Bi**	209	499544.2	1.5				ug/L	523984	Standard

Sample ID: L1606066107
Report Date/Time: Wednesday, June 29, 2016 12:17:43
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**90.1824**	29.248	32.4	mg/L	0	Standard
Mg	24	323.3	10.9	**0.3294**	0.041	12.6	mg/L	37	Standard
K	39	788.4	10.3	**1.5298**	0.110	7.2	mg/L	38	Standard
Ca	43	663.3	6.3	**37.0621**	1.711	4.6	mg/L	10	Standard
Fe	54	18630.5	1.2	**14.8018**	0.561	3.8	mg/L	179	Standard
Fe	57	5110.9	3.8	**15.3724**	1.217	7.9	mg/L	335	Standard
Sc-1	45	64585.8	4.1				mg/L	61029	Standard
Cl	35	22235.6	2.9				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	23104.3	4.2				ug/L	1547	Standard
P	31	15829.9	2.1				ug/L	15286	Standard
S	34	2033.5	3.2				ug/L	1572	Standard
Sr	88	161.7	12.5				ug/L	73	Standard
C	12	96.7	41.8				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	31177.2	1.6				mg/L	6	Standard
Ho-1	165	18616.4	1.9				mg/L	10	Standard
Er	166	16976.2	1.7				mg/L	13	Standard
I	127	9242.8	4.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		106.903	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.682	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606066107
Report Date/Time: Wednesday, June 29, 2016 12:17:43
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.350
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.336
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606066107
Report Date/Time: Wednesday, June 29, 2016 12:17:43
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066108
Sample Date/Time: Wednesday, June 29, 2016 12:18:38
Number of Replicates: 3
Autosampler Position: 303
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77986.9	1.5				ug/L	74699	Standard
	Be	9	848.4	8.3	**0.9313**	0.087	9.4	ug/L	8	Standard
	Al	27	17300686.4	1.1	**112.7510**	0.475	0.4	ug/L	1493	Standard
	Sc	45	64965.7	2.0				ug/L	61029	Standard
	Ti	47	6993.9	0.6	**18.3803**	0.198	1.1	ug/L	35	Standard
	V	51	170039.3	0.3	**22.3624**	0.095	0.4	ug/L	1435	Standard
	Cr	52	135852.7	1.1	**17.3429**	0.173	1.0	ug/L	7635	Standard
	Cr	53	16400.5	1.1	**17.4496**	0.150	0.9	ug/L	363	Standard
	Mn	55	3732645.2	0.8	**537.0232**	5.661	1.1	ug/L	747	Standard
	Co	59	59499.9	1.1	**8.9910**	0.146	1.6	ug/L	234	Standard
	Ni	60	38977.4	1.0	**20.7345**	0.185	0.9	ug/L	215	Standard
	Cu	65	32324.4	0.5	**17.2976**	0.098	0.6	ug/L	104	Standard
	Zn	66	55484.7	0.8	**50.4392**	0.445	0.9	ug/L	363	Standard
>	**Ge**	72	494184.7	0.5				ug/L	495142	Standard
	As	75	5935.1	0.6	**5.3688**	0.057	1.1	ug/L	-50	Standard
	Se	82	56.5	10.0	**0.3782**	0.052	13.7	ug/L	14	Standard
	Se-1	77	118.0	0.8	**0.9965**	0.019	1.9	ug/L	42	Standard
>	**Ga**	71	16687.5	1.9				mg/L	42	Standard
	Rb	85	157851.4	1.2				ug/L	17	Standard
	Y	89	635672.9	1.6				ug/L	417098	Standard
>	**Rh**	103	68.3	27.7				ug/L	13	Standard
	Mo	98	1214.1	0.7	**0.4777**	0.002	0.4	ug/L	11	Standard
	Ag	107	767.0	3.0	**0.1120**	0.004	3.7	ug/L	68	Standard
	Cd	111	562.6	3.0	**0.2776**	0.008	3.0	mg/L	2	Standard
	Cd	114	1264.6	12.4	**0.2596**	0.032	12.4	ug/L	12	Standard
>	**In**	115	553515.9	0.3				ug/L	587441	Standard
	Sn	118	763.4	11.0	**0.0440**	0.015	34.7	ug/L	497	Standard
	Sb	123	242.9	15.0	**0.0510**	0.008	16.4	ug/L	78	Standard
	Ba	135	429181.6	1.6	**222.5293**	3.912	1.8	ug/L	24	Standard
	Ce	140	1102655.1	2.3				ug/L	25	Standard
>	**Tb**	159	923824.1	1.4				ug/L	939701	Standard
	Ho	165	18663.2	2.9				ug/L	10	Standard
	Tl	203	782.7	3.6	**0.0981**	0.005	5.3	ug/L	4	Standard
	Tl	205	701.7	8.6	**0.1034**	0.010	10.1	ug/L	2	Standard
	Pb	206	213507.8	0.5	**41.9534**	0.898	2.1	ug/L	240	Standard
	Pb	207	175202.2	0.5	**38.7416**	0.683	1.8	ug/L	198	Standard
	Pb	208	727362.3	0.6	**40.7140**	1.033	2.5	ug/L	883	Standard
	U	238	4175.6	1.0	**0.6335**	0.018	2.9	ug/L	3	Standard
>	**Bi**	209	498479.4	1.9				ug/L	523984	Standard

Sample ID: L1606066108
Report Date/Time: Wednesday, June 29, 2016 12:20:55
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.0364**	31.231	173.2	mg/L	0	Standard
Mg	24	140.0	21.7	**0.1047**	0.037	35.2	mg/L	37	Standard
K	39	708.4	12.8	**1.3550**	0.168	12.4	mg/L	38	Standard
Ca	43	508.3	5.0	**27.6291**	0.895	3.2	mg/L	10	Standard
Fe	54	15324.4	1.1	**12.0707**	0.309	2.6	mg/L	179	Standard
Fe	57	4262.3	3.3	**12.5338**	0.279	2.2	mg/L	335	Standard
> **Sc-1**	45	64965.7	2.0				mg/L	61029	Standard
Cl	35	21927.9	1.6				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	2823.6	3.3				ug/L	1547	Standard
P	31	16612.4	1.0				ug/L	15286	Standard
S	34	2085.1	7.7				ug/L	1572	Standard
Sr	88	245.0	14.1				ug/L	73	Standard
C	12	116.7	27.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	93.3	49.5				mg/L	3	Standard
Dy	164	30954.0	1.2				mg/L	6	Standard
Ho-1	165	18663.2	2.9				mg/L	10	Standard
Er	166	16615.8	2.6				mg/L	13	Standard
I	127	11766.2	5.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		104.402	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		99.807	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606066108
Report Date/Time: Wednesday, June 29, 2016 12:20:55
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	94.225	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	95.132	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606066108
Report Date/Time: Wednesday, June 29, 2016 12:20:55
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066109
Sample Date/Time: Wednesday, June 29, 2016 12:21:49
Number of Replicates: 3
Autosampler Position: 304
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77037.0	1.5				ug/L	74699	Standard
	Be	9	378.3	6.2	**0.4242**	0.022	5.2	ug/L	8	Standard
	Al	27	13254869.5	1.3	**87.4672**	2.449	2.8	ug/L	1493	Standard
	Sc	45	63133.0	3.0				ug/L	61029	Standard
	Ti	47	8144.2	1.7	**21.2616**	0.854	4.0	ug/L	35	Standard
	V	51	109381.5	1.5	**14.2029**	0.301	2.1	ug/L	1435	Standard
	Cr	52	127693.7	2.3	**16.1132**	0.755	4.7	ug/L	7635	Standard
	Cr	53	15149.2	3.9	**15.9774**	1.000	6.3	ug/L	363	Standard
	Mn	55	3010101.1	0.9	**429.7771**	13.412	3.1	ug/L	747	Standard
	Co	59	32158.1	0.7	**4.8045**	0.098	2.0	ug/L	234	Standard
	Ni	60	23283.5	0.8	**12.2424**	0.383	3.1	ug/L	215	Standard
	Cu	65	9940.9	0.4	**5.2071**	0.131	2.5	ug/L	104	Standard
	Zn	66	24785.3	0.2	**22.1054**	0.470	2.1	ug/L	363	Standard
>	**Ge**	72	498160.0	2.3				ug/L	495142	Standard
	As	75	2943.4	0.7	**2.6727**	0.077	2.9	ug/L	-50	Standard
	Se	82	39.1	13.2	**0.2170**	0.054	24.7	ug/L	14	Standard
	Se-1	77	75.7	11.2	**0.4068**	0.103	25.3	ug/L	42	Standard
>	**Ga**	71	9034.4	2.0				mg/L	42	Standard
	Rb	85	68682.0	2.5				ug/L	17	Standard
	Y	89	539979.1	1.9				ug/L	417098	Standard
>	**Rh**	103	11.7	24.7				ug/L	13	Standard
	Mo	98	3058.1	1.5	**1.1809**	0.018	1.5	ug/L	11	Standard
	Ag	107	334.3	34.2	**0.0400**	0.017	43.3	ug/L	68	Standard
	Cd	111	137.9	39.4	**0.0601**	0.026	43.1	mg/L	2	Standard
	Cd	114	340.6	39.5	**0.0664**	0.026	39.4	ug/L	12	Standard
>	**In**	115	568023.6	1.3				ug/L	587441	Standard
	Sn	118	973.4	3.3	**0.0776**	0.003	4.4	ug/L	497	Standard
	Sb	123	217.2	33.7	**0.0438**	0.016	35.6	ug/L	78	Standard
	Ba	135	138215.3	1.3	**69.8220**	0.106	0.2	ug/L	24	Standard
	Ce	140	545871.5	1.0				ug/L	25	Standard
>	**Tb**	159	946980.1	0.3				ug/L	939701	Standard
	Ho	165	11547.7	1.8				ug/L	10	Standard
	Tl	203	458.7	30.8	**0.0539**	0.017	31.9	ug/L	4	Standard
	Tl	205	360.0	14.1	**0.0505**	0.007	14.1	ug/L	2	Standard
	Pb	206	30575.5	1.0	**5.6549**	0.063	1.1	ug/L	240	Standard
	Pb	207	24590.3	1.3	**5.1132**	0.071	1.4	ug/L	198	Standard
	Pb	208	101567.1	0.7	**5.3447**	0.038	0.7	ug/L	883	Standard
	U	238	5591.0	1.3	**0.8082**	0.011	1.4	ug/L	3	Standard
>	**Bi**	209	523773.4	0.1				ug/L	523984	Standard

Sample ID: L1606066109
Report Date/Time: Wednesday, June 29, 2016 12:24:06
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.1528**	33.165	173.2	mg/L	0	Standard
Mg	24	88.3	26.1	**0.0447**	0.026	57.5	mg/L	37	Standard
K	39	410.0	24.0	**0.7638**	0.217	28.4	mg/L	38	Standard
Ca	43	368.3	13.4	**20.0310**	3.643	18.2	mg/L	10	Standard
Fe	54	11096.4	1.9	**8.9581**	0.101	1.1	mg/L	179	Standard
Fe	57	2985.3	3.6	**8.7334**	0.110	1.3	mg/L	335	Standard
Sc-1	45	63133.0	3.0				mg/L	61029	Standard
Cl	35	21859.8	1.0				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	1696.8	5.1				ug/L	1547	Standard
P	31	15684.8	2.6				ug/L	15286	Standard
S	34	1883.5	4.2				ug/L	1572	Standard
Sr	88	116.7	19.3				ug/L	73	Standard
C	12	100.0	26.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	18768.7	3.5				mg/L	6	Standard
Ho-1	165	11547.7	1.8				mg/L	10	Standard
Er	166	10790.5	1.6				mg/L	13	Standard
I	127	6703.2	5.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		103.130	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		100.609	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606066109
Report Date/Time: Wednesday, June 29, 2016 12:24:06
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.695
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.960
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606066109
Report Date/Time: Wednesday, June 29, 2016 12:24:06
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082203
Sample Date/Time: Wednesday, June 29, 2016 12:25:01
Number of Replicates: 3
Autosampler Position: 305
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75246.5	2.8				ug/L	74699	Standard
	Be	9	6.7	114.6	**0.0150**	0.008	55.4	ug/L	8	Standard
	Al	27	1140453.8	1.3	**7.6935**	0.186	2.4	ug/L	1493	Standard
	Sc	45	58880.9	1.2				ug/L	61029	Standard
	Ti	47	77.7	5.2	**0.0982**	0.010	10.0	ug/L	35	Standard
	V	51	1502.7	7.0	**0.0058**	0.017	290.5	ug/L	1435	Standard
	Cr	52	10404.9	0.3	**0.3964**	0.023	5.7	ug/L	7635	Standard
	Cr	53	878.4	11.3	**0.6323**	0.120	19.0	ug/L	363	Standard
	Mn	55	64288.8	1.0	**9.1920**	0.174	1.9	ug/L	747	Standard
	Co	59	1249.1	3.1	**0.1540**	0.006	4.1	ug/L	234	Standard
	Ni	60	673.3	4.1	**0.2398**	0.011	4.6	ug/L	215	Standard
	Cu	65	229.7	8.0	**0.0211**	0.010	48.0	ug/L	104	Standard
	Zn	66	2146.2	1.5	**1.5328**	0.038	2.5	ug/L	363	Standard
>	**Ge**	72	489582.8	1.5				ug/L	495142	Standard
	As	75	99.3	9.7	**0.1492**	0.008	5.6	ug/L	-50	Standard
	Se	82	43.5	22.3	**0.2617**	0.085	32.3	ug/L	14	Standard
	Se-1	77	67.7	6.2	**0.3148**	0.059	18.8	ug/L	42	Standard
>	**Ga**	71	61.7	40.8				mg/L	42	Standard
	Rb	85	88606.7	2.0				ug/L	17	Standard
	Y	89	430024.7	1.6				ug/L	417098	Standard
>	**Rh**	103	28.3	50.9				ug/L	13	Standard
	Mo	98	81.5	16.2	**0.0249**	0.005	20.1	ug/L	11	Standard
	Ag	107	97.7	18.1	**0.0021**	0.003	124.3	ug/L	68	Standard
	Cd	111	13.9	57.3	**-0.0013**	0.004	291.1	mg/L	2	Standard
	Cd	114	23.0	35.9	**0.0024**	0.002	67.9	ug/L	12	Standard
>	**In**	115	586341.4	0.7				ug/L	587441	Standard
	Sn	118	508.3	13.9	**-0.0076**	0.011	151.7	ug/L	497	Standard
	Sb	123	156.7	24.2	**0.0294**	0.008	27.7	ug/L	78	Standard
	Ba	135	76489.9	1.2	**37.4252**	0.302	0.8	ug/L	24	Standard
	Ce	140	665.0	13.3				ug/L	25	Standard
>	**Tb**	159	934158.4	1.1				ug/L	939701	Standard
	Ho	165	15.0	66.7				ug/L	10	Standard
	Tl	203	113.7	11.3	**0.0121**	0.002	13.7	ug/L	4	Standard
	Tl	205	93.3	44.6	**0.0133**	0.006	44.3	ug/L	2	Standard
	Pb	206	362.0	5.5	**-0.0029**	0.004	131.1	ug/L	240	Standard
	Pb	207	281.3	6.6	**-0.0103**	0.004	38.7	ug/L	198	Standard
	Pb	208	1197.0	7.1	**-0.0098**	0.005	49.2	ug/L	883	Standard
	U	238	25.0	34.9	**-0.0008**	0.001	168.5	ug/L	3	Standard
>	**Bi**	209	520173.1	0.7				ug/L	523984	Standard

Sample ID: L1606082203
Report Date/Time: Wednesday, June 29, 2016 12:27:18
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	3547.1	5.1	**4.6793**	0.186	4.0	mg/L	37	Standard
K	39	1096.7	1.1	**2.3987**	0.055	2.3	mg/L	38	Standard
Ca	43	51.7	27.9	**0.8562**	0.931	108.7	mg/L	10	Standard
Fe	54	433.8	1.2	**0.2430**	0.009	3.5	mg/L	179	Standard
Fe	57	376.7	13.8	**0.2532**	0.189	74.5	mg/L	335	Standard
> **Sc-1**	45	58880.9	1.2				mg/L	61029	Standard
Cl	35	22906.6	0.4				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	9496.3	3.0				ug/L	1547	Standard
P	31	16589.1	2.9				ug/L	15286	Standard
S	34	2071.8	1.6				ug/L	1572	Standard
Sr	88	118.3	17.1				ug/L	73	Standard
C	12	53.3	10.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	19.0	47.5				mg/L	6	Standard
Ho-1	165	15.0	66.7				mg/L	10	Standard
Er	166	20.0	100.0				mg/L	13	Standard
I	127	14315.1	5.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		100.733	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.877	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082203
Report Date/Time: Wednesday, June 29, 2016 12:27:18
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.813
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.273
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082302
Sample Date/Time: Wednesday, June 29, 2016 12:28:12
Number of Replicates: 3
Autosampler Position: 306
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75794.1	0.6				ug/L	74699	Standard
	Be	9	13.3	21.7	**0.0225**	0.003	14.2	ug/L	8	Standard
	Al	27	2805598.0	0.7	**18.8025**	0.242	1.3	ug/L	1493	Standard
	Sc	45	59848.0	1.2				ug/L	61029	Standard
	Ti	47	42.3	20.1	**0.0017**	0.021	1272.3	ug/L	35	Standard
	V	51	1764.1	8.5	**0.0361**	0.018	49.9	ug/L	1435	Standard
	Cr	52	8282.9	3.5	**0.0853**	0.037	43.4	ug/L	7635	Standard
	Cr	53	906.7	14.2	**0.6437**	0.138	21.5	ug/L	363	Standard
	Mn	55	61274.7	1.5	**8.5848**	0.068	0.8	ug/L	747	Standard
	Co	59	3907.8	2.1	**0.5500**	0.015	2.7	ug/L	234	Standard
	Ni	60	654.7	3.9	**0.2231**	0.012	5.6	ug/L	215	Standard
	Cu	65	222.3	1.8	**0.0148**	0.001	8.1	ug/L	104	Standard
	Zn	66	2204.5	3.0	**1.5481**	0.054	3.5	ug/L	363	Standard
>	**Ge**	72	498979.3	0.8				ug/L	495142	Standard
	As	75	72.7	46.8	**0.1239**	0.030	24.0	ug/L	-50	Standard
	Se	82	52.0	6.1	**0.3317**	0.029	8.6	ug/L	14	Standard
	Se-1	77	105.0	8.3	**0.8037**	0.108	13.5	ug/L	42	Standard
>	**Ga**	71	56.7	13.5				mg/L	42	Standard
	Rb	85	343.3	5.1				ug/L	17	Standard
	Y	89	428520.0	1.9				ug/L	417098	Standard
>	**Rh**	103	38.3	27.2				ug/L	13	Standard
	Mo	98	38.4	23.3	**0.0089**	0.004	39.5	ug/L	11	Standard
	Ag	107	77.7	2.7	**-0.0008**	0.000	47.2	ug/L	68	Standard
	Cd	111	5.6	57.3	**-0.0052**	0.002	30.1	mg/L	2	Standard
	Cd	114	18.7	97.5	**0.0016**	0.004	227.0	ug/L	12	Standard
>	**In**	115	580343.3	0.7				ug/L	587441	Standard
	Sn	118	485.0	23.2	**-0.0106**	0.020	190.8	ug/L	497	Standard
	Sb	123	114.5	31.8	**0.0207**	0.008	39.1	ug/L	78	Standard
	Ba	135	150889.3	1.2	**74.6061**	0.512	0.7	ug/L	24	Standard
	Ce	140	123.3	15.3				ug/L	25	Standard
>	**Tb**	159	936316.1	1.2				ug/L	939701	Standard
	Ho	165	18.3	31.5				ug/L	10	Standard
	Tl	203	189.0	9.5	**0.0214**	0.002	8.8	ug/L	4	Standard
	Tl	205	161.7	6.4	**0.0230**	0.002	7.6	ug/L	2	Standard
	Pb	206	327.3	6.0	**-0.0093**	0.004	47.2	ug/L	240	Standard
	Pb	207	269.3	8.1	**-0.0127**	0.005	36.6	ug/L	198	Standard
	Pb	208	1102.3	4.7	**-0.0147**	0.003	21.5	ug/L	883	Standard
	U	238	62.7	10.4	**0.0048**	0.001	17.9	ug/L	3	Standard
>	**Bi**	209	519166.0	1.5				ug/L	523984	Standard

Sample ID: L1606082302
Report Date/Time: Wednesday, June 29, 2016 12:30:29
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	5325.9	0.2	**6.9463**	0.099	1.4	mg/L	37	Standard
K	39	75.0	37.1	**0.0544**	0.061	111.8	mg/L	38	Standard
Ca	43	86.7	3.3	**3.0583**	0.130	4.3	mg/L	10	Standard
Fe	54	170.8	11.8	**0.0094**	0.018	186.9	mg/L	179	Standard
Fe	57	311.7	10.3	**0.0068**	0.121	1775.6	mg/L	335	Standard
> **Sc-1**	45	59848.0	1.2				mg/L	61029	Standard
Cl	35	23747.3	1.3				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	11654.5	4.8				ug/L	1547	Standard
P	31	16051.8	4.0				ug/L	15286	Standard
S	34	1918.5	1.8				ug/L	1572	Standard
Sr	88	140.0	9.4				ug/L	73	Standard
C	12	53.3	39.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	12.2	186.8				mg/L	6	Standard
Ho-1	165	18.3	31.5				mg/L	10	Standard
Er	166	23.3	24.7				mg/L	13	Standard
I	127	13899.7	5.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.466	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.775	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082302
Report Date/Time: Wednesday, June 29, 2016 12:30:29
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.792
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.080
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082303
Sample Date/Time: Wednesday, June 29, 2016 12:31:24
Number of Replicates: 3
Autosampler Position: 307
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76244.7	1.6				ug/L	74699	Standard
	Be	9	10.0	50.0	**0.0187**	0.005	28.8	ug/L	8	Standard
	Al	27	1789441.8	2.2	**11.9153**	0.077	0.6	ug/L	1493	Standard
	Sc	45	61370.7	1.9				ug/L	61029	Standard
	Ti	47	59.7	11.8	**0.0448**	0.018	40.0	ug/L	35	Standard
	V	51	1976.3	1.1	**0.0601**	0.005	8.2	ug/L	1435	Standard
	Cr	52	8828.6	1.1	**0.1403**	0.016	11.5	ug/L	7635	Standard
	Cr	53	945.0	0.5	**0.6692**	0.006	1.0	ug/L	363	Standard
	Mn	55	180374.5	2.7	**25.1617**	0.439	1.7	ug/L	747	Standard
	Co	59	619.0	3.1	**0.0543**	0.004	7.0	ug/L	234	Standard
	Ni	60	405.7	6.0	**0.0878**	0.011	12.8	ug/L	215	Standard
	Cu	65	208.0	10.3	**0.0056**	0.012	222.0	ug/L	104	Standard
	Zn	66	6989.3	1.3	**5.7979**	0.130	2.2	ug/L	363	Standard
>	**Ge**	72	506777.4	1.1				ug/L	495142	Standard
	As	75	17.4	166.6	**0.0747**	0.025	33.6	ug/L	-50	Standard
	Se	82	43.0	5.9	**0.2447**	0.022	8.9	ug/L	14	Standard
	Se-1	77	89.0	9.0	**0.5677**	0.097	17.0	ug/L	42	Standard
>	**Ga**	71	55.0	18.2				mg/L	42	Standard
	Rb	85	226.7	20.8				ug/L	17	Standard
	Y	89	430230.4	1.6				ug/L	417098	Standard
>	**Rh**	103	23.3	12.4				ug/L	13	Standard
	Mo	98	43.4	8.2	**0.0107**	0.002	14.5	ug/L	11	Standard
	Ag	107	80.3	9.2	**-0.0005**	0.001	248.5	ug/L	68	Standard
	Cd	111	11.9	16.8	**-0.0022**	0.001	43.2	mg/L	2	Standard
	Cd	114	28.1	50.1	**0.0034**	0.003	82.8	ug/L	12	Standard
>	**In**	115	584282.7	1.3				ug/L	587441	Standard
	Sn	118	445.0	7.0	**-0.0181**	0.005	24.8	ug/L	497	Standard
	Sb	123	113.1	24.2	**0.0201**	0.006	28.5	ug/L	78	Standard
	Ba	135	83408.3	1.0	**40.9596**	0.601	1.5	ug/L	24	Standard
	Ce	140	301.7	22.6				ug/L	25	Standard
>	**Tb**	159	946714.8	0.6				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	182.7	3.3	**0.0205**	0.001	3.9	ug/L	4	Standard
	Tl	205	160.0	30.8	**0.0227**	0.007	30.9	ug/L	2	Standard
	Pb	206	288.0	1.3	**-0.0170**	0.001	5.3	ug/L	240	Standard
	Pb	207	270.7	6.0	**-0.0126**	0.003	25.4	ug/L	198	Standard
	Pb	208	1097.0	10.5	**-0.0153**	0.006	38.6	ug/L	883	Standard
	U	238	866.7	1.0	**0.1222**	0.001	0.8	ug/L	3	Standard
>	**Bi**	209	520921.7	0.5				ug/L	523984	Standard

Sample ID: L1606082303
Report Date/Time: Wednesday, June 29, 2016 12:33:40
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.3655**	33.533	173.2	mg/L	0	Standard
Mg	24	4915.8	1.0	**6.2466**	0.131	2.1	mg/L	37	Standard
K	39	50.0	20.0	**-0.0042**	0.023	541.6	mg/L	38	Standard
Ca	43	65.0	0.0	**1.5606**	0.079	5.0	mg/L	10	Standard
Fe	54	160.0	11.0	**-0.0035**	0.013	359.3	mg/L	179	Standard
Fe	57	281.7	9.1	**-0.1230**	0.068	55.3	mg/L	335	Standard
> **Sc-1**	45	61370.7	1.9				mg/L	61029	Standard
Cl	35	23872.8	0.3				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	8675.8	8.1				ug/L	1547	Standard
P	31	16263.7	3.0				ug/L	15286	Standard
S	34	1945.1	6.6				ug/L	1572	Standard
Sr	88	130.0	24.0				ug/L	73	Standard
C	12	46.7	44.6				mg/L	63	Standard
N	14	6.7	86.6				mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	19.0	57.5				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	20.0	100.0				mg/L	13	Standard
I	127	19577.6	2.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.070	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.350	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082303
Report Date/Time: Wednesday, June 29, 2016 12:33:40
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.462
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.415
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082304
Sample Date/Time: Wednesday, June 29, 2016 12:34:35
Number of Replicates: 3
Autosampler Position: 308
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76966.7	2.1				ug/L	74699	Standard
	Be	9	8.3	91.7	**0.0169**	0.009	50.5	ug/L	8	Standard
	Al	27	441491.6	0.4	**2.9030**	0.060	2.1	ug/L	1493	Standard
	Sc	45	60155.9	2.7				ug/L	61029	Standard
	Ti	47	44.0	12.0	**0.0058**	0.013	219.4	ug/L	35	Standard
	V	51	1675.2	9.7	**0.0239**	0.019	79.8	ug/L	1435	Standard
	Cr	52	8137.2	4.7	**0.0630**	0.039	62.0	ug/L	7635	Standard
	Cr	53	611.7	11.7	**0.3248**	0.069	21.3	ug/L	363	Standard
	Mn	55	105254.7	1.8	**14.8182**	0.156	1.1	ug/L	747	Standard
	Co	59	259.7	5.8	**0.0017**	0.003	156.1	ug/L	234	Standard
	Ni	60	230.0	4.9	**-0.0022**	0.007	321.8	ug/L	215	Standard
	Cu	65	165.3	1.5	**-0.0157**	0.002	12.9	ug/L	104	Standard
	Zn	66	2212.5	1.5	**1.5507**	0.010	0.6	ug/L	363	Standard
>	**Ge**	72	500132.0	1.1				ug/L	495142	Standard
	As	75	21.8	38.4	**0.0789**	0.008	9.6	ug/L	-50	Standard
	Se	82	26.9	8.6	**0.1047**	0.023	22.2	ug/L	14	Standard
	Se-1	77	64.0	26.1	**0.2460**	0.230	93.4	ug/L	42	Standard
>	**Ga**	71	55.0	41.7				mg/L	42	Standard
	Rb	85	421.7	18.5				ug/L	17	Standard
	Y	89	429897.2	1.0				ug/L	417098	Standard
>	**Rh**	103	6.7	43.3				ug/L	13	Standard
	Mo	98	29.2	13.0	**0.0052**	0.001	25.4	ug/L	11	Standard
	Ag	107	87.7	6.3	**0.0005**	0.001	182.2	ug/L	68	Standard
	Cd	111	7.0	28.7	**-0.0046**	0.001	19.4	mg/L	2	Standard
	Cd	114	21.6	74.7	**0.0021**	0.003	150.8	ug/L	12	Standard
>	**In**	115	589054.0	1.8				ug/L	587441	Standard
	Sn	118	438.3	5.7	**-0.0199**	0.003	16.6	ug/L	497	Standard
	Sb	123	106.2	8.9	**0.0185**	0.002	12.5	ug/L	78	Standard
	Ba	135	25689.8	1.2	**12.5057**	0.379	3.0	ug/L	24	Standard
	Ce	140	65.0	26.6				ug/L	25	Standard
>	**Tb**	159	953435.2	1.7				ug/L	939701	Standard
	Ho	165	13.3	43.3				ug/L	10	Standard
	Tl	203	91.7	10.9	**0.0091**	0.001	12.8	ug/L	4	Standard
	Tl	205	101.7	11.4	**0.0141**	0.002	11.0	ug/L	2	Standard
	Pb	206	281.0	2.2	**-0.0197**	0.001	5.9	ug/L	240	Standard
	Pb	207	219.0	9.1	**-0.0248**	0.004	16.3	ug/L	198	Standard
	Pb	208	999.0	3.0	**-0.0219**	0.001	6.3	ug/L	883	Standard
	U	238	2.7	21.7	**-0.0040**	0.000	2.0	ug/L	3	Standard
>	**Bi**	209	535313.4	0.3				ug/L	523984	Standard

Sample ID: L1606082304
Report Date/Time: Wednesday, June 29, 2016 12:36:52
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	1101.7	8.8	**1.3766**	0.099	7.2	mg/L	37	Standard
K	39	50.0	26.5	**-0.0017**	0.031	1783.7	mg/L	38	Standard
Ca	43	43.3	17.6	**0.2622**	0.555	211.4	mg/L	10	Standard
Fe	54	1079.1	4.5	**0.7903**	0.045	5.7	mg/L	179	Standard
Fe	57	550.0	4.8	**0.8245**	0.134	16.2	mg/L	335	Standard
> **Sc-1**	45	60155.9	2.7				mg/L	61029	Standard
Cl	35	23751.9	0.9				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	3913.8	0.4				ug/L	1547	Standard
P	31	16552.4	4.0				ug/L	15286	Standard
S	34	1948.5	5.9				ug/L	1572	Standard
Sr	88	68.3	30.5				ug/L	73	Standard
C	12	63.3	36.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	19.7	135.8				mg/L	6	Standard
Ho-1	165	13.3	43.3				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	5207.6	8.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.036	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.008	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082304
Report Date/Time: Wednesday, June 29, 2016 12:36:52
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.275
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.162
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 12:37:48
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72653.8	1.6				ug/L	74699	Standard
	Be	9	42693.4	2.4	**49.8778**	1.581	3.2	ug/L	8	Standard
	Al	27	7018269.3	1.5	**49.0877**	0.211	0.4	ug/L	1493	Standard
	Sc	45	58700.2	1.7				ug/L	61029	Standard
	Ti	47	37360.7	1.5	**96.4941**	1.122	1.2	ug/L	35	Standard
	V	51	389361.0	0.6	**50.3345**	0.950	1.9	ug/L	1435	Standard
	Cr	52	377308.6	0.5	**48.8761**	0.835	1.7	ug/L	7635	Standard
	Cr	53	46063.3	3.3	**48.5325**	2.229	4.6	ug/L	363	Standard
	Mn	55	355991.7	2.4	**49.9570**	0.746	1.5	ug/L	747	Standard
	Co	59	341523.2	1.7	**50.6452**	0.528	1.0	ug/L	234	Standard
	Ni	60	94080.5	0.6	**49.1230**	0.490	1.0	ug/L	215	Standard
	Cu	65	95078.4	0.9	**49.9689**	1.178	2.4	ug/L	104	Standard
	Zn	66	56040.5	0.8	**49.8274**	0.722	1.4	ug/L	363	Standard
>	**Ge**	72	505267.2	1.6				ug/L	495142	Standard
	As	75	56871.7	1.3	**49.8175**	0.372	0.7	ug/L	-50	Standard
	Se	82	5593.2	2.1	**49.8004**	0.347	0.7	ug/L	14	Standard
	Se-1	77	3775.1	1.7	**50.0311**	0.568	1.1	ug/L	42	Standard
>	**Ga**	71	75.0	13.3				mg/L	42	Standard
	Rb	85	800.0	10.3				ug/L	17	Standard
	Y	89	408700.7	0.8				ug/L	417098	Standard
>	**Rh**	103	18.3	41.7				ug/L	13	Standard
	Mo	98	257910.1	1.9	**100.4657**	4.126	4.1	ug/L	11	Standard
	Ag	107	316989.6	1.3	**50.4692**	0.906	1.8	ug/L	68	Standard
	Cd	111	101914.0	1.0	**50.5948**	1.545	3.1	mg/L	2	Standard
	Cd	114	242699.9	2.4	**49.1311**	2.013	4.1	ug/L	12	Standard
>	**In**	115	566154.7	2.7				ug/L	587441	Standard
	Sn	118	281464.1	1.7	**49.9153**	1.113	2.2	ug/L	497	Standard
	Sb	123	227316.6	0.8	**50.5420**	1.534	3.0	ug/L	78	Standard
	Ba	135	98379.4	0.9	**49.8785**	1.233	2.5	ug/L	24	Standard
	Ce	140	113.3	20.8				ug/L	25	Standard
>	**Tb**	159	932106.1	1.1				ug/L	939701	Standard
	Ho	165	8.3	91.7				ug/L	10	Standard
	Tl	203	406761.4	0.1	**49.7003**	0.911	1.8	ug/L	4	Standard
	Tl	205	354916.5	2.0	**49.9771**	1.877	3.8	ug/L	2	Standard
	Pb	206	263562.6	0.6	**49.6234**	0.874	1.8	ug/L	240	Standard
	Pb	207	235278.4	0.8	**49.8612**	1.016	2.0	ug/L	198	Standard
	Pb	208	929662.3	0.8	**49.8673**	1.247	2.5	ug/L	883	Standard
	U	238	336021.4	1.5	**49.1730**	1.579	3.2	ug/L	3	Standard
>	**Bi**	209	520331.8	1.7				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:40:05
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.5945**	33.930	173.2	mg/L	0	Standard
Mg	24	3858.8	4.6	**5.1144**	0.246	4.8	mg/L	37	Standard
K	39	2055.1	3.6	**4.6098**	0.203	4.4	mg/L	38	Standard
Ca	43	130.0	23.4	**5.9979**	1.898	31.6	mg/L	10	Standard
Fe	54	5737.9	3.9	**4.9227**	0.278	5.7	mg/L	179	Standard
Fe	57	1583.4	3.8	**4.5245**	0.301	6.7	mg/L	335	Standard
Sc-1	45	58700.2	1.7				mg/L	61029	Standard
Cl	35	23099.6	1.9				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	1760.1	10.8				ug/L	1547	Standard
P	31	17002.9	0.9				ug/L	15286	Standard
S	34	2305.2	8.6				ug/L	1572	Standard
Sr	88	76.7	10.0				ug/L	73	Standard
C	12	60.0	33.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	9.0	191.6				mg/L	6	Standard
Ho-1	165	8.3	91.7				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	2030.1	31.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	99.756		
Al	27	98.175		
Sc	45			
Ti	47	96.494		
V	51	100.669		
Cr	52	97.752		
Cr	53			
Mn	55	99.914		
Co	59	101.290		
Ni	60	98.246		
Cu	65	99.938		
Zn	66	99.655		
Ge	72		102.045	
As	75	99.635		
Se	82	99.601		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:40:05
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	100.466		
Ag	107	100.938		
Cd	111	101.190		
Cd	114			
> In	115			96.376
Sn	118	99.831		
Sb	123	101.084		
Ba	135	99.757		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	99.401		
Tl	205			
Pb	206			
Pb	207			
Pb	208	99.735		
U	238	98.346		
> Bi	209			99.303
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:40:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 12:40:59
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73745.7	1.0				ug/L	74699	Standard
	Be	9	45.0	86.8	**0.0596**	0.045	76.2	ug/L	8	Standard
	Al	27	10763.7	138.2	**0.0612**	0.103	168.8	ug/L	1493	Standard
	Sc	45	59794.4	0.5				ug/L	61029	Standard
	Ti	47	63.7	40.2	**0.0565**	0.070	123.9	ug/L	35	Standard
	V	51	1309.0	5.1	**-0.0256**	0.009	36.2	ug/L	1435	Standard
	Cr	52	7144.0	4.7	**-0.0780**	0.064	81.5	ug/L	7635	Standard
	Cr	53	391.7	5.2	**0.0859**	0.030	34.5	ug/L	363	Standard
	Mn	55	2203.3	114.3	**0.1667**	0.364	218.6	ug/L	747	Standard
	Co	59	375.0	36.8	**0.0187**	0.022	116.6	ug/L	234	Standard
	Ni	60	226.7	25.3	**-0.0048**	0.033	679.9	ug/L	215	Standard
	Cu	65	137.7	45.7	**-0.0307**	0.035	113.4	ug/L	104	Standard
	Zn	66	391.7	8.9	**-0.1025**	0.038	37.5	ug/L	363	Standard
>	**Ge**	72	505094.9	1.9				ug/L	495142	Standard
	As	75	-30.0	135.0	**0.0336**	0.035	105.0	ug/L	-50	Standard
	Se	82	18.1	21.6	**0.0240**	0.035	145.0	ug/L	14	Standard
	Se-1	77	51.7	16.7	**0.0700**	0.104	148.9	ug/L	42	Standard
>	**Ga**	71	55.0	27.3				mg/L	42	Standard
	Rb	85	30.0	72.6				ug/L	17	Standard
	Y	89	416389.1	0.5				ug/L	417098	Standard
>	**Rh**	103	18.3	56.8				ug/L	13	Standard
	Mo	98	292.6	24.5	**0.1060**	0.029	27.7	ug/L	11	Standard
	Ag	107	123.7	36.8	**0.0063**	0.007	108.0	ug/L	68	Standard
	Cd	111	19.1	60.5	**0.0013**	0.005	411.5	mg/L	2	Standard
	Cd	114	51.4	75.2	**0.0080**	0.007	93.3	ug/L	12	Standard
>	**In**	115	578912.4	1.7				ug/L	587441	Standard
	Sn	118	756.7	23.6	**0.0364**	0.029	79.3	ug/L	497	Standard
	Sb	123	391.4	30.7	**0.0806**	0.025	30.9	ug/L	78	Standard
	Ba	135	64.3	74.2	**0.0157**	0.023	147.1	ug/L	24	Standard
	Ce	140	98.3	142.4				ug/L	25	Standard
>	**Tb**	159	938659.9	2.1				ug/L	939701	Standard
	Ho	165	16.7	45.8				ug/L	10	Standard
	Tl	203	80.3	33.4	**0.0079**	0.003	39.1	ug/L	4	Standard
	Tl	205	263.3	142.0	**0.0362**	0.051	140.6	ug/L	2	Standard
	Pb	206	290.0	20.9	**-0.0175**	0.011	60.1	ug/L	240	Standard
	Pb	207	255.0	4.1	**-0.0168**	0.002	14.3	ug/L	198	Standard
	Pb	208	1806.4	80.0	**0.0207**	0.074	360.4	ug/L	883	Standard
	U	238	234.3	149.2	**0.0289**	0.049	171.5	ug/L	3	Standard
>	**Bi**	209	529175.1	1.6				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:43:16
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	48.3	47.8	**-0.0014**	0.030	2108.5	mg/L	37	Standard
K	39	68.3	15.2	**0.0400**	0.023	56.9	mg/L	38	Standard
Ca	43	21.7	74.2	**-1.1227**	1.047	93.2	mg/L	10	Standard
Fe	54	169.2	23.1	**0.0081**	0.034	414.8	mg/L	179	Standard
Fe	57	310.0	16.4	**0.0019**	0.182	9789.6	mg/L	335	Standard
Sc-1	45	59794.4	0.5				mg/L	61029	Standard
Cl	35	23378.7	0.4				ug/L	20195	Standard
Kr	83	4.3	58.1				ug/L	4	Standard
Br	81	1630.1	3.1				ug/L	1547	Standard
P	31	16767.6	2.6				ug/L	15286	Standard
S	34	2153.5	8.8				ug/L	1572	Standard
Sr	88	65.0	27.7				ug/L	73	Standard
C	12	66.7	17.3				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	9.4	186.5				mg/L	6	Standard
Ho-1	165	16.7	45.8				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	1680.1	29.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		102.010	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:43:16
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	98.548	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	100.991	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:43:16
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Wednesday, June 29, 2016 12:44:12
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	71218.7	0.8				ug/L	74699	Standard
	Be	9	178.3	13.3	**0.2201**	0.029	13.2	ug/L	8	Standard
	Al	27	1766.8	29.5	**-0.0008**	0.004	458.8	ug/L	1493	Standard
	Sc	45	56025.1	1.1				ug/L	61029	Standard
	Ti	47	28.0	7.1	**-0.0318**	0.007	20.5	ug/L	35	Standard
	V	51	4079.9	5.1	**0.3696**	0.041	11.2	ug/L	1435	Standard
	Cr	52	12699.0	0.7	**0.7676**	0.041	5.4	ug/L	7635	Standard
	Cr	53	1050.0	6.6	**0.8579**	0.085	9.9	ug/L	363	Standard
	Mn	55	4110.2	0.9	**0.4692**	0.021	4.5	ug/L	747	Standard
	Co	59	2790.3	0.6	**0.4045**	0.014	3.5	ug/L	234	Standard
	Ni	60	2962.3	1.6	**1.5299**	0.023	1.5	ug/L	215	Standard
	Cu	65	1521.1	1.5	**0.7509**	0.034	4.5	ug/L	104	Standard
	Zn	66	7672.3	2.1	**6.8861**	0.035	0.5	ug/L	363	Standard
>	Ge	72	473802.8	2.6				ug/L	495142	Standard
	As	75	358.5	4.9	**0.3943**	0.021	5.3	ug/L	-50	Standard
	Se	82	48.3	16.1	**0.3230**	0.086	26.5	ug/L	14	Standard
	Se-1	77	78.0	9.0	**0.4927**	0.075	15.2	ug/L	42	Standard
>	Ga	71	35.0	24.7				mg/L	42	Standard
	Rb	85	21.7	35.3				ug/L	17	Standard
	Y	89	406616.3	2.6				ug/L	417098	Standard
>	Rh	103	13.3	43.3				ug/L	13	Standard
	Mo	98	80.8	10.7	**0.0260**	0.003	11.1	ug/L	11	Standard
	Ag	107	2461.5	2.7	**0.3827**	0.004	1.0	ug/L	68	Standard
	Cd	111	469.2	3.6	**0.2272**	0.006	2.5	mg/L	2	Standard
	Cd	114	1121.7	10.1	**0.2273**	0.026	11.5	ug/L	12	Standard
>	In	115	560717.1	1.7				ug/L	587441	Standard
	Sn	118	496.7	12.6	**-0.0055**	0.012	217.4	ug/L	497	Standard
	Sb	123	1804.7	2.2	**0.4010**	0.016	3.9	ug/L	78	Standard
	Ba	135	1433.1	2.6	**0.7176**	0.020	2.8	ug/L	24	Standard
	Ce	140	53.3	43.3				ug/L	25	Standard
>	Tb	159	901191.9	2.3				ug/L	939701	Standard
	Ho	165	5.0	0.0				ug/L	10	Standard
	Tl	203	658.3	6.7	**0.0809**	0.005	6.0	ug/L	4	Standard
	Tl	205	580.0	6.7	**0.0841**	0.005	5.7	ug/L	2	Standard
	Pb	206	1314.1	4.2	**0.1836**	0.008	4.1	ug/L	240	Standard
	Pb	207	1090.4	5.0	**0.1679**	0.009	5.1	ug/L	198	Standard
	Pb	208	4577.3	1.3	**0.1788**	0.005	3.0	ug/L	883	Standard
	U	238	2625.2	2.7	**0.3906**	0.007	1.9	ug/L	3	Standard
>	Bi	209	505831.0	1.4				ug/L	523984	Standard

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 12:46:29
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	40.0	12.5	**-0.0088**	0.007	78.9	mg/L	37	Standard
K	39	38.3	19.9	**-0.0218**	0.018	84.6	mg/L	38	Standard
Ca	43	36.7	75.1	**-0.0119**	1.877	15817.5	mg/L	10	Standard
Fe	54	97.1	24.3	**-0.0484**	0.023	47.3	mg/L	179	Standard
Fe	57	295.0	11.9	**0.0175**	0.123	699.1	mg/L	335	Standard
Sc-1	45	56025.1	1.1				mg/L	61029	Standard
Cl	35	22767.8	1.3				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	1303.4	5.0				ug/L	1547	Standard
P	31	11956.4	1.3				ug/L	15286	Standard
S	34	2061.8	6.1				ug/L	1572	Standard
Sr	88	80.0	22.5				ug/L	73	Standard
C	12	80.0	33.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	0.0					mg/L	6	Standard
Ho-1	165	5.0	0.0				mg/L	10	Standard
Er	166	0.0					mg/L	13	Standard
I	127	1360.1	35.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	110.075		
Al	27			
Sc	45			
Ti	47			
V	51	92.400		
Cr	52	95.955		
Cr	53			
Mn	55	93.844		
Co	59	101.134		
Ni	60	95.616		
Cu	65	93.867		
Zn	66	110.178		
Ge	72		95.690	
As	75	98.586		
Se	82	80.738		
Se-1	77			
Ga	71			

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 12:46:29
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	95.685		
Cd	111	94.646		
Cd	114			
> In	115			95.451
Sn	118			
Sb	123	100.239		
Ba	135	95.683		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.186		
Tl	205			
Pb	206			
Pb	207			
Pb	208	89.410		
U	238	97.655		
> Bi	209			96.535
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 12:46:29
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBW 7 WG573548-02
Sample Date/Time: Wednesday, June 29, 2016 12:53:54
Number of Replicates: 3
Autosampler Position: 309
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74857.8	1.6				ug/L	74699	Standard
	Be	9	16.7	75.5	**0.0264**	0.014	53.3	ug/L	8	Standard
	Al	27	80794.9	2.0	**0.5353**	0.014	2.5	ug/L	1493	Standard
	Sc	45	58461.0	3.0				ug/L	61029	Standard
	Ti	47	78.0	5.9	**0.0921**	0.013	14.5	ug/L	35	Standard
	V	51	1804.8	0.8	**0.0380**	0.004	11.0	ug/L	1435	Standard
	Cr	52	9792.5	0.9	**0.2676**	0.021	7.7	ug/L	7635	Standard
	Cr	53	633.3	5.7	**0.3397**	0.038	11.1	ug/L	363	Standard
	Mn	55	7868.0	0.9	**0.9563**	0.019	2.0	ug/L	747	Standard
	Co	59	389.7	6.7	**0.0204**	0.003	16.2	ug/L	234	Standard
	Ni	60	396.3	4.2	**0.0829**	0.008	9.4	ug/L	215	Standard
	Cu	65	182.0	8.2	**-0.0081**	0.009	109.2	ug/L	104	Standard
	Zn	66	1432.4	5.9	**0.8274**	0.087	10.5	ug/L	363	Standard
>	**Ge**	72	506709.7	1.0				ug/L	495142	Standard
	As	75	61.7	54.0	**0.1136**	0.029	25.8	ug/L	-50	Standard
	Se	82	57.7	7.0	**0.3762**	0.040	10.6	ug/L	14	Standard
	Se-1	77	48.3	5.2	**0.0243**	0.031	128.3	ug/L	42	Standard
>	**Ga**	71	106.7	35.5				mg/L	42	Standard
	Rb	85	413.3	4.9				ug/L	17	Standard
	Y	89	414741.4	0.4				ug/L	417098	Standard
>	**Rh**	103	15.0	100.0				ug/L	13	Standard
	Mo	98	46.5	23.9	**0.0125**	0.004	33.5	ug/L	11	Standard
	Ag	107	1549.7	3.3	**0.2354**	0.012	5.1	ug/L	68	Standard
	Cd	111	17.2	8.9	**0.0007**	0.001	113.7	mg/L	2	Standard
	Cd	114	63.6	36.2	**0.0109**	0.005	44.4	ug/L	12	Standard
>	**In**	115	562869.1	1.7				ug/L	587441	Standard
	Sn	118	736.7	11.6	**0.0370**	0.017	44.6	ug/L	497	Standard
	Sb	123	216.9	10.9	**0.0443**	0.006	12.6	ug/L	78	Standard
	Ba	135	906.0	5.8	**0.4461**	0.030	6.6	ug/L	24	Standard
	Ce	140	1735.1	5.3				ug/L	25	Standard
>	**Tb**	159	918530.3	1.8				ug/L	939701	Standard
	Ho	165	31.7	32.9				ug/L	10	Standard
	Tl	203	89.3	15.5	**0.0092**	0.002	18.3	ug/L	4	Standard
	Tl	205	86.7	13.3	**0.0124**	0.002	13.0	ug/L	2	Standard
	Pb	206	378.0	4.1	**0.0001**	0.003	2503.4	ug/L	240	Standard
	Pb	207	310.7	3.7	**-0.0040**	0.002	56.8	ug/L	198	Standard
	Pb	208	1268.7	1.9	**-0.0059**	0.001	22.1	ug/L	883	Standard
	U	238	52.0	18.9	**0.0032**	0.001	44.2	ug/L	3	Standard
>	**Bi**	209	520017.2	0.7				ug/L	523984	Standard

Sample ID: PBW 7 WG573548-02
Report Date/Time: Wednesday, June 29, 2016 12:56:11
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	98.3	32.7	**0.0678**	0.045	67.0	mg/L	37	Standard
K	39	61.7	9.4	**0.0281**	0.012	42.4	mg/L	38	Standard
Ca	43	31.7	55.5	**-0.4154**	1.202	289.4	mg/L	10	Standard
Fe	54	153.7	8.4	**-0.0024**	0.007	308.5	mg/L	179	Standard
Fe	57	316.7	15.8	**0.0490**	0.173	353.8	mg/L	335	Standard
> **Sc-1**	45	58461.0	3.0				mg/L	61029	Standard
Cl	35	22819.8	1.0				ug/L	20195	Standard
Kr	83	3.7	31.5				ug/L	4	Standard
Br	81	12281.6	4.1				ug/L	1547	Standard
P	31	15094.2	2.5				ug/L	15286	Standard
S	34	1958.5	7.0				ug/L	1572	Standard
Sr	88	78.3	19.5				ug/L	73	Standard
C	12	96.7	21.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	48.1	42.1				mg/L	6	Standard
Ho-1	165	31.7	32.9				mg/L	10	Standard
Er	166	40.0	25.0				mg/L	13	Standard
I	127	2060.1	27.3				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		100.213	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		102.336	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBW 7 WG573548-02
Report Date/Time: Wednesday, June 29, 2016 12:56:11
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.817
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.243
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 7 WG573548-02
Report Date/Time: Wednesday, June 29, 2016 12:56:11
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSW 7 WG573548-03
Sample Date/Time: Wednesday, June 29, 2016 12:57:05
Number of Replicates: 3
Autosampler Position: 310
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75779.1	1.4				ug/L	74699	Standard
	Be	9	44266.3	2.0	**49.5806**	1.413	2.8	ug/L	8	Standard
	Al	27	95277.1	2.2	**0.6259**	0.022	3.5	ug/L	1493	Standard
	Sc	45	60008.6	1.7				ug/L	61029	Standard
	Ti	47	61.3	19.5	**0.0533**	0.033	61.1	ug/L	35	Standard
	V	51	383994.5	1.5	**50.7641**	0.334	0.7	ug/L	1435	Standard
	Cr	52	391642.5	1.8	**51.9470**	0.996	1.9	ug/L	7635	Standard
	Cr	53	48176.5	1.3	**51.9174**	0.408	0.8	ug/L	363	Standard
	Mn	55	379234.4	0.1	**54.4505**	0.406	0.7	ug/L	747	Standard
	Co	59	337441.4	0.9	**51.1841**	0.650	1.3	ug/L	234	Standard
	Ni	60	96944.2	1.6	**51.7745**	0.652	1.3	ug/L	215	Standard
	Cu	65	97447.4	0.9	**52.3759**	0.338	0.6	ug/L	104	Standard
	Zn	66	58710.9	1.4	**53.4195**	0.683	1.3	ug/L	363	Standard
>	**Ge**	72	493989.7	0.8				ug/L	495142	Standard
	As	75	54420.0	2.0	**48.7560**	0.745	1.5	ug/L	-50	Standard
	Se	82	5499.8	2.1	**50.0913**	1.110	2.2	ug/L	14	Standard
	Se-1	77	3757.5	2.3	**50.9460**	1.277	2.5	ug/L	42	Standard
>	**Ga**	71	331.7	6.3				mg/L	42	Standard
	Rb	85	1905.1	7.7				ug/L	17	Standard
	Y	89	430193.5	2.2				ug/L	417098	Standard
>	**Rh**	103	21.7	13.3				ug/L	13	Standard
	Mo	98	200.3	5.0	**0.0686**	0.004	5.6	ug/L	11	Standard
	Ag	107	325181.2	0.5	**49.3255**	0.456	0.9	ug/L	68	Standard
	Cd	111	102382.7	0.7	**48.4138**	0.589	1.2	mg/L	2	Standard
	Cd	114	246010.3	1.2	**47.4297**	0.453	1.0	ug/L	12	Standard
>	**In**	115	594099.2	0.9				ug/L	587441	Standard
	Sn	118	733.4	7.5	**0.0295**	0.010	35.4	ug/L	497	Standard
	Sb	123	222790.7	0.6	**47.1842**	0.681	1.4	ug/L	78	Standard
	Ba	135	100456.3	0.7	**48.5175**	0.574	1.2	ug/L	24	Standard
	Ce	140	3297.0	2.9				ug/L	25	Standard
>	**Tb**	159	953975.0	2.3				ug/L	939701	Standard
	Ho	165	81.7	3.5				ug/L	10	Standard
	Tl	203	411383.7	1.0	**49.0691**	0.983	2.0	ug/L	4	Standard
	Tl	205	358822.0	2.1	**49.3118**	1.192	2.4	ug/L	2	Standard
	Pb	206	272876.4	0.6	**50.1549**	0.805	1.6	ug/L	240	Standard
	Pb	207	232281.8	1.2	**48.0517**	0.976	2.0	ug/L	198	Standard
	Pb	208	938885.2	1.2	**49.1601**	1.120	2.3	ug/L	883	Standard
	U	238	337957.9	0.3	**48.2703**	0.691	1.4	ug/L	3	Standard
>	**Bi**	209	532979.3	1.2				ug/L	523984	Standard

Sample ID: LCSW 7 WG573548-03
Report Date/Time: Wednesday, June 29, 2016 12:59:22
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	40.0	12.5	**-0.0126**	0.007	52.7	mg/L	37	Standard
K	39	58.3	21.6	**0.0173**	0.030	174.9	mg/L	38	Standard
Ca	43	60.0	16.7	**1.3313**	0.636	47.7	mg/L	10	Standard
Fe	54	427.3	5.7	**0.2300**	0.018	7.6	mg/L	179	Standard
Fe	57	393.3	4.8	**0.2864**	0.089	31.1	mg/L	335	Standard
> **Sc-1**	45	60008.6	1.7				mg/L	61029	Standard
Cl	35	22904.0	1.4				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	1933.5	11.4				ug/L	1547	Standard
P	31	15693.1	2.0				ug/L	15286	Standard
S	34	1913.5	5.4				ug/L	1572	Standard
Sr	88	85.0	21.2				ug/L	73	Standard
C	12	106.7	37.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	149.3	43.4				mg/L	6	Standard
Ho-1	165	81.7	3.5				mg/L	10	Standard
Er	166	83.3	45.4				mg/L	13	Standard
I	127	1690.1	32.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.446	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.767	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 7 WG573548-03
Report Date/Time: Wednesday, June 29, 2016 12:59:22
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.133
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.717
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 7 WG573548-03
Report Date/Time: Wednesday, June 29, 2016 12:59:22
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098623 WG573548-01
Sample Date/Time: Wednesday, June 29, 2016 13:00:17
Number of Replicates: 3
Autosampler Position: 311
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85536.1	4.6				ug/L	74699	Standard
	Be	9	131.7	4.4	**0.1385**	0.009	6.7	ug/L	8	Standard
	Al	27	51229429.5	1.9	**304.6734**	8.454	2.8	ug/L	1493	Standard
	Sc	45	58156.4	0.7				ug/L	61029	Standard
	Ti	47	291.7	4.0	**0.6976**	0.027	3.9	ug/L	35	Standard
	V	51	2731.2	8.5	**0.1837**	0.035	19.3	ug/L	1435	Standard
	Cr	52	11798.6	2.2	**0.6453**	0.034	5.3	ug/L	7635	Standard
	Cr	53	1541.7	10.7	**1.4199**	0.200	14.1	ug/L	363	Standard
	Mn	55	1519614.4	1.3	**228.6645**	1.850	0.8	ug/L	747	Standard
	Co	59	1958.8	0.9	**0.2737**	0.000	0.1	ug/L	234	Standard
	Ni	60	5664.7	1.6	**3.0483**	0.071	2.3	ug/L	215	Standard
	Cu	65	19925.4	0.1	**11.1201**	0.082	0.7	ug/L	104	Standard
	Zn	66	6263.0	2.6	**5.5559**	0.126	2.3	ug/L	363	Standard
>	**Ge**	72	472304.7	0.8				ug/L	495142	Standard
	As	75	721.9	3.2	**0.7353**	0.022	3.1	ug/L	-50	Standard
	Se	82	400.4	3.0	**3.6870**	0.142	3.9	ug/L	14	Standard
	Se-1	77	187.0	7.2	**2.0610**	0.173	8.4	ug/L	42	Standard
>	**Ga**	71	100.0	8.7				mg/L	42	Standard
	Rb	85	5192.6	2.9				ug/L	17	Standard
	Y	89	424150.6	2.5				ug/L	417098	Standard
>	**Rh**	103	291.7	21.8				ug/L	13	Standard
	Mo	98	1037.1	3.0	**0.3904**	0.016	4.1	ug/L	11	Standard
	Ag	107	150.3	11.8	**0.0106**	0.003	25.5	ug/L	68	Standard
	Cd	111	86.5	9.0	**0.0342**	0.004	11.5	mg/L	2	Standard
	Cd	114	235.1	23.0	**0.0445**	0.010	23.2	ug/L	12	Standard
>	**In**	115	577173.5	1.1				ug/L	587441	Standard
	Sn	118	1115.0	2.8	**0.0996**	0.004	4.3	ug/L	497	Standard
	Sb	123	1725.9	23.2	**0.3715**	0.084	22.5	ug/L	78	Standard
	Ba	135	22344.1	0.8	**11.0965**	0.201	1.8	ug/L	24	Standard
	Ce	140	1543.4	4.9				ug/L	25	Standard
>	**Tb**	159	936227.5	1.2				ug/L	939701	Standard
	Ho	165	78.3	35.2				ug/L	10	Standard
	Tl	203	240.0	4.8	**0.0287**	0.001	5.1	ug/L	4	Standard
	Tl	205	171.7	9.4	**0.0253**	0.002	9.3	ug/L	2	Standard
	Pb	206	666.7	6.9	**0.0592**	0.009	15.2	ug/L	240	Standard
	Pb	207	558.7	0.1	**0.0529**	0.000	0.3	ug/L	198	Standard
	Pb	208	2267.7	0.3	**0.0522**	0.000	0.9	ug/L	883	Standard
	U	238	28225.3	0.9	**4.2785**	0.042	1.0	ug/L	3	Standard
>	**Bi**	209	501665.1	0.1				ug/L	523984	Standard

Sample ID: L1606098623 WG573548-01
Report Date/Time: Wednesday, June 29, 2016 13:02:34
Page 1

Approved: July 04, 2016

Na	23	13.3	43.3	**161.5718**	71.131	44.0	mg/L	0	Standard
Mg	24	24900.4	1.1	**33.6647**	0.149	0.4	mg/L	37	Standard
K	39	1233.4	1.4	**2.7469**	0.062	2.2	mg/L	38	Standard
Ca	43	1560.1	5.5	**100.8648**	4.911	4.9	mg/L	10	Standard
Fe	54	167.1	25.8	**0.0105**	0.039	372.7	mg/L	179	Standard
Fe	57	958.4	6.3	**2.3440**	0.212	9.0	mg/L	335	Standard
Sc-1	45	58156.4	0.7				mg/L	61029	Standard
Cl	35	40465.3	2.5				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	58241.9	4.1				ug/L	1547	Standard
P	31	17453.4	1.6				ug/L	15286	Standard
S	34	1810.1	8.9				ug/L	1572	Standard
Sr	88	1313.4	8.1				ug/L	73	Standard
C	12	273.3	15.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	3	Standard
Dy	164	89.3	16.3				mg/L	6	Standard
Ho-1	165	78.3	35.2				mg/L	10	Standard
Er	166	83.3	18.3				mg/L	13	Standard
I	127	135078.2	7.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		114.508	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.388	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098623 WG573548-01
Report Date/Time: Wednesday, June 29, 2016 13:02:34
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.252
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.740
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098635S WG573548-04
Sample Date/Time: Wednesday, June 29, 2016 13:03:28
Number of Replicates: 3
Autosampler Position: 312
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96645.6	0.2				ug/L	74699	Standard
	Be	9	56332.9	2.6	**49.4619**	1.222	2.5	ug/L	8	Standard
	Al	27	56565131.5	1.0	**297.4859**	3.609	1.2	ug/L	1493	Standard
	Sc	45	58999.8	4.0				ug/L	61029	Standard
	Ti	47	369.0	11.5	**0.8968**	0.108	12.0	ug/L	35	Standard
	V	51	390242.6	0.5	**53.2266**	0.984	1.8	ug/L	1435	Standard
	Cr	52	371857.5	2.1	**50.8419**	0.408	0.8	ug/L	7635	Standard
	Cr	53	46974.4	2.3	**52.2114**	1.040	2.0	ug/L	363	Standard
	Mn	55	1851169.7	0.3	**274.7367**	3.654	1.3	ug/L	747	Standard
	Co	59	350457.1	0.5	**54.8326**	0.988	1.8	ug/L	234	Standard
	Ni	60	94372.4	2.0	**51.9777**	0.274	0.5	ug/L	215	Standard
	Cu	65	115580.4	2.6	**64.0828**	0.734	1.1	ug/L	104	Standard
	Zn	66	61118.4	0.5	**57.3902**	0.612	1.1	ug/L	363	Standard
>	**Ge**	72	478993.7	1.6				ug/L	495142	Standard
	As	75	59122.1	0.8	**54.6260**	0.454	0.8	ug/L	-50	Standard
	Se	82	6294.5	0.9	**59.1515**	0.578	1.0	ug/L	14	Standard
	Se-1	77	4101.2	2.7	**57.4162**	0.662	1.2	ug/L	42	Standard
>	**Ga**	71	100.0	18.0				mg/L	42	Standard
	Rb	85	5364.3	1.3				ug/L	17	Standard
	Y	89	423648.2	2.3				ug/L	417098	Standard
>	**Rh**	103	290.0	15.8				ug/L	13	Standard
	Mo	98	1231.1	6.2	**0.4537**	0.027	5.8	ug/L	11	Standard
	Ag	107	140053.7	1.3	**21.3673**	0.341	1.6	ug/L	68	Standard
	Cd	111	101619.5	1.1	**48.3438**	0.154	0.3	mg/L	2	Standard
	Cd	114	248812.6	0.3	**48.2667**	0.577	1.2	ug/L	12	Standard
>	**In**	115	590497.7	1.3				ug/L	587441	Standard
	Sn	118	1276.7	2.9	**0.1227**	0.007	5.6	ug/L	497	Standard
	Sb	123	230648.9	1.4	**49.1452**	0.703	1.4	ug/L	78	Standard
	Ba	135	121100.4	1.0	**58.8551**	1.255	2.1	ug/L	24	Standard
	Ce	140	1580.1	12.1				ug/L	25	Standard
>	**Tb**	159	967679.2	1.8				ug/L	939701	Standard
	Ho	165	70.0	12.4				ug/L	10	Standard
	Tl	203	411213.8	0.9	**50.0486**	0.144	0.3	ug/L	4	Standard
	Tl	205	371375.3	1.6	**52.0881**	1.342	2.6	ug/L	2	Standard
	Pb	206	271184.0	0.3	**50.8655**	0.565	1.1	ug/L	240	Standard
	Pb	207	230172.1	0.3	**48.5918**	0.740	1.5	ug/L	198	Standard
	Pb	208	934549.0	0.8	**49.9358**	0.874	1.8	ug/L	883	Standard
	U	238	386476.6	1.1	**56.3336**	1.024	1.8	ug/L	3	Standard
>	**Bi**	209	522269.6	1.2				ug/L	523984	Standard

Sample ID: L1606098635S WG573548-04
Report Date/Time: Wednesday, June 29, 2016 13:05:45
Page 1

Approved: July 04, 2016

Na	23	8.3	91.7	**99.7813**	89.869	90.1	mg/L	0	Standard
Mg	24	26446.4	1.6	**35.2703**	0.863	2.4	mg/L	37	Standard
K	39	1201.7	4.6	**2.6333**	0.063	2.4	mg/L	38	Standard
Ca	43	1648.4	4.4	**105.1755**	1.490	1.4	mg/L	10	Standard
Fe	54	207.2	35.2	**0.0431**	0.061	140.9	mg/L	179	Standard
Fe	57	976.7	6.7	**2.3580**	0.131	5.6	mg/L	335	Standard
> **Sc-1**	45	58999.8	4.0				mg/L	61029	Standard
Cl	35	42664.0	0.9				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	62512.1	2.5				ug/L	1547	Standard
P	31	18094.1	2.0				ug/L	15286	Standard
S	34	1855.1	3.1				ug/L	1572	Standard
Sr	88	1376.7	4.2				ug/L	73	Standard
C	12	206.7	24.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	73.5	28.0				mg/L	6	Standard
Ho-1	165	70.0	12.4				mg/L	10	Standard
Er	166	66.7	34.6				mg/L	13	Standard
I	127	146421.6	4.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		129.380	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.739	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098635S WG573548-04
Report Date/Time: Wednesday, June 29, 2016 13:05:45
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	100.520
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	99.673
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098637SD WG573548-05
Sample Date/Time: Wednesday, June 29, 2016 13:06:39
Number of Replicates: 3
Autosampler Position: 313
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	99478.9	1.8				ug/L	74699	Standard
	Be	9	55809.3	1.7	**47.6121**	0.679	1.4	ug/L	8	Standard
	Al	27	57079276.5	2.7	**291.6498**	6.801	2.3	ug/L	1493	Standard
	Sc	45	56220.8	1.7				ug/L	61029	Standard
	Ti	47	357.7	29.4	**0.8728**	0.310	35.5	ug/L	35	Standard
	V	51	392795.1	0.3	**53.7087**	1.385	2.6	ug/L	1435	Standard
	Cr	52	364211.2	1.2	**49.9063**	1.054	2.1	ug/L	7635	Standard
	Cr	53	45092.0	2.6	**50.2220**	0.967	1.9	ug/L	363	Standard
	Mn	55	1864825.2	1.1	**277.4410**	6.534	2.4	ug/L	747	Standard
	Co	59	357216.6	1.3	**56.0166**	0.665	1.2	ug/L	234	Standard
	Ni	60	94064.4	0.7	**51.9434**	1.099	2.1	ug/L	215	Standard
	Cu	65	96872.1	1.3	**53.8290**	0.581	1.1	ug/L	104	Standard
	Zn	66	56756.1	1.1	**53.3877**	0.715	1.3	ug/L	363	Standard
>	**Ge**	72	477920.9	2.4				ug/L	495142	Standard
	As	75	58834.8	0.8	**54.4924**	0.918	1.7	ug/L	-50	Standard
	Se	82	6411.7	0.6	**60.4029**	1.057	1.7	ug/L	14	Standard
	Se-1	77	4156.6	3.0	**58.3446**	1.540	2.6	ug/L	42	Standard
>	**Ga**	71	126.7	16.4				mg/L	42	Standard
	Rb	85	5344.3	1.7				ug/L	17	Standard
	Y	89	425761.1	0.9				ug/L	417098	Standard
>	**Rh**	103	338.3	14.4				ug/L	13	Standard
	Mo	98	943.7	3.1	**0.3564**	0.013	3.7	ug/L	11	Standard
	Ag	107	126054.2	0.9	**19.7696**	0.296	1.5	ug/L	68	Standard
	Cd	111	100507.1	1.6	**49.1528**	0.361	0.7	mg/L	2	Standard
	Cd	114	240725.5	3.1	**47.9933**	0.862	1.8	ug/L	12	Standard
>	**In**	115	574417.4	1.5				ug/L	587441	Standard
	Sn	118	1058.4	4.2	**0.0905**	0.005	5.8	ug/L	497	Standard
	Sb	123	229849.5	0.5	**50.3475**	0.456	0.9	ug/L	78	Standard
	Ba	135	119199.1	0.7	**59.5474**	0.534	0.9	ug/L	24	Standard
	Ce	140	1411.7	10.5				ug/L	25	Standard
>	**Tb**	159	946510.5	1.0				ug/L	939701	Standard
	Ho	165	61.7	32.8				ug/L	10	Standard
	Tl	203	404314.5	1.0	**50.2968**	0.720	1.4	ug/L	4	Standard
	Tl	205	368989.1	0.6	**52.8952**	1.157	2.2	ug/L	2	Standard
	Pb	206	274418.5	1.5	**52.6043**	0.331	0.6	ug/L	240	Standard
	Pb	207	233408.5	1.0	**50.3596**	0.394	0.8	ug/L	198	Standard
	Pb	208	945417.2	0.6	**51.6312**	0.704	1.4	ug/L	883	Standard
	U	238	390395.6	1.2	**58.1569**	0.854	1.5	ug/L	3	Standard
>	**Bi**	209	511024.4	1.7				ug/L	523984	Standard

Sample ID: L1606098637SD WG573548-05
Report Date/Time: Wednesday, June 29, 2016 13:08:56
Page 1

Approved: July 04, 2016

Na	23	13.3	21.7	**167.0389**	37.375	22.4	mg/L	0	Standard
Mg	24	26820.4	2.0	**37.5161**	0.222	0.6	mg/L	37	Standard
K	39	1165.0	5.8	**2.6799**	0.118	4.4	mg/L	38	Standard
Ca	43	1491.7	10.9	**99.7885**	11.256	11.3	mg/L	10	Standard
Fe	54	157.1	11.5	**0.0062**	0.015	238.1	mg/L	179	Standard
Fe	57	1011.7	5.7	**2.6573**	0.164	6.2	mg/L	335	Standard
> **Sc-1**	45	56220.8	1.7				mg/L	61029	Standard
Cl	35	42892.6	1.5				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	62984.0	1.3				ug/L	1547	Standard
P	31	17081.3	3.0				ug/L	15286	Standard
S	34	1805.1	2.7				ug/L	1572	Standard
Sr	88	1383.4	10.5				ug/L	73	Standard
C	12	180.0	14.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	124.6	33.0				mg/L	6	Standard
Ho-1	165	61.7	32.8				mg/L	10	Standard
Er	166	43.3	35.3				mg/L	13	Standard
I	127	140810.4	4.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		133.173	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.522	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098637SD WG573548-05
Report Date/Time: Wednesday, June 29, 2016 13:08:56
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.783
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	97.527
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098637SD WG573548-05
Report Date/Time: Wednesday, June 29, 2016 13:08:56
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098601
Sample Date/Time: Wednesday, June 29, 2016 13:09:51
Number of Replicates: 3
Autosampler Position: 314
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	102324.4	1.8				ug/L	74699	Standard
	Be	9	88.3	32.2	**0.0810**	0.024	30.0	ug/L	8	Standard
	Al	27	5114148.7	1.9	**25.3952**	0.605	2.4	ug/L	1493	Standard
	Sc	45	60191.0	1.9				ug/L	61029	Standard
	Ti	47	2779.9	0.6	**7.0936**	0.108	1.5	ug/L	35	Standard
	V	51	5111.4	5.6	**0.4688**	0.033	7.0	ug/L	1435	Standard
	Cr	52	13420.3	3.5	**0.7542**	0.051	6.8	ug/L	7635	Standard
	Cr	53	1458.4	3.3	**1.2196**	0.051	4.2	ug/L	363	Standard
	Mn	55	181173.8	0.3	**25.4046**	0.180	0.7	ug/L	747	Standard
	Co	59	2348.5	2.0	**0.3119**	0.007	2.1	ug/L	234	Standard
	Ni	60	2310.5	3.1	**1.0878**	0.031	2.9	ug/L	215	Standard
	Cu	65	2308.5	3.6	**1.1141**	0.034	3.0	ug/L	104	Standard
	Zn	66	22662.3	0.5	**19.9177**	0.090	0.4	ug/L	363	Standard
>	**Ge**	72	504254.4	0.9				ug/L	495142	Standard
	As	75	859.1	5.1	**0.8126**	0.036	4.5	ug/L	-50	Standard
	Se	82	56.8	14.9	**0.3708**	0.079	21.2	ug/L	14	Standard
	Se-1	77	87.3	17.2	**0.5507**	0.190	34.5	ug/L	42	Standard
>	**Ga**	71	526.7	8.3				mg/L	42	Standard
	Rb	85	16517.3	5.3				ug/L	17	Standard
	Y	89	430302.4	1.6				ug/L	417098	Standard
>	**Rh**	103	86.7	33.8				ug/L	13	Standard
	Mo	98	1898.6	1.4	**0.7136**	0.022	3.1	ug/L	11	Standard
	Ag	107	115.3	8.0	**0.0050**	0.001	29.0	ug/L	68	Standard
	Cd	111	60.8	33.2	**0.0215**	0.010	47.6	mg/L	2	Standard
	Cd	114	128.6	20.8	**0.0232**	0.005	20.9	ug/L	12	Standard
>	**In**	115	581907.5	1.7				ug/L	587441	Standard
	Sn	118	1231.7	10.8	**0.1183**	0.024	20.7	ug/L	497	Standard
	Sb	123	520.8	14.7	**0.1084**	0.016	15.1	ug/L	78	Standard
	Ba	135	41346.0	0.9	**20.3791**	0.217	1.1	ug/L	24	Standard
	Ce	140	23020.8	2.4				ug/L	25	Standard
>	**Tb**	159	964144.3	1.8				ug/L	939701	Standard
	Ho	165	513.3	9.3				ug/L	10	Standard
	Tl	203	211.3	12.7	**0.0234**	0.003	14.7	ug/L	4	Standard
	Tl	205	253.3	62.3	**0.0350**	0.022	63.0	ug/L	2	Standard
	Pb	206	1500.1	2.6	**0.2039**	0.008	4.0	ug/L	240	Standard
	Pb	207	1190.0	4.0	**0.1756**	0.010	5.7	ug/L	198	Standard
	Pb	208	4936.0	2.0	**0.1838**	0.008	4.3	ug/L	883	Standard
	U	238	2034.5	10.8	**0.2852**	0.034	11.8	ug/L	3	Standard
>	**Bi**	209	535012.0	1.1				ug/L	523984	Standard

Sample ID: L1606098601
Report Date/Time: Wednesday, June 29, 2016 13:12:08
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	50263.2	0.4	**65.7389**	1.477	2.2	mg/L	37	Standard
K	39	250.0	7.2	**0.4466**	0.047	10.5	mg/L	38	Standard
Ca	43	233.3	8.9	**12.4104**	1.127	9.1	mg/L	10	Standard
Fe	54	825.5	3.8	**0.5718**	0.038	6.7	mg/L	179	Standard
Fe	57	516.7	4.4	**0.7058**	0.045	6.4	mg/L	335	Standard
> **Sc-1**	45	60191.0	1.9				mg/L	61029	Standard
Cl	35	28383.9	2.7				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	8439.0	2.4				ug/L	1547	Standard
P	31	17857.2	2.8				ug/L	15286	Standard
S	34	1791.8	4.6				ug/L	1572	Standard
Sr	88	506.7	4.7				ug/L	73	Standard
C	12	153.3	22.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	893.4	9.1				mg/L	6	Standard
Ho-1	165	513.3	9.3				mg/L	10	Standard
Er	166	416.7	13.2				mg/L	13	Standard
I	127	18778.3	7.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		136.983	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.840	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098601
Report Date/Time: Wednesday, June 29, 2016 13:12:08
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	99.058	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	102.105	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603
Sample Date/Time: Wednesday, June 29, 2016 13:13:02
Number of Replicates: 3
Autosampler Position: 315
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	106961.3	1.6				ug/L	74699	Standard
	Be	9	178.3	7.1	**0.1492**	0.012	8.2	ug/L	8	Standard
	Al	27	64490531.4	4.4	**306.3886**	9.864	3.2	ug/L	1493	Standard
	Sc	45	59811.2	2.1				ug/L	61029	Standard
	Ti	47	416.0	0.2	**0.9910**	0.025	2.5	ug/L	35	Standard
	V	51	3004.3	2.1	**0.2029**	0.003	1.2	ug/L	1435	Standard
	Cr	52	11901.7	1.5	**0.5854**	0.019	3.3	ug/L	7635	Standard
	Cr	53	991.7	2.3	**0.7459**	0.050	6.7	ug/L	363	Standard
	Mn	55	12370472.2	4.2	**1779.3602**	31.656	1.8	ug/L	747	Standard
	Co	59	18875.4	5.7	**2.8249**	0.100	3.5	ug/L	234	Standard
	Ni	60	12202.3	4.0	**6.4037**	0.114	1.8	ug/L	215	Standard
	Cu	65	16836.0	2.8	**8.9582**	0.039	0.4	ug/L	104	Standard
	Zn	66	24869.4	2.3	**22.3549**	0.039	0.2	ug/L	363	Standard
>	**Ge**	72	494237.2	2.4				ug/L	495142	Standard
	As	75	1554.4	6.7	**1.4489**	0.061	4.2	ug/L	-50	Standard
	Se	82	71.5	7.7	**0.5142**	0.037	7.1	ug/L	14	Standard
	Se-1	77	63.0	12.4	**0.2431**	0.119	49.1	ug/L	42	Standard
>	**Ga**	71	356.7	10.3				mg/L	42	Standard
	Rb	85	9309.5	2.8				ug/L	17	Standard
	Y	89	460268.8	4.1				ug/L	417098	Standard
>	**Rh**	103	310.0	9.7				ug/L	13	Standard
	Mo	98	1407.8	3.5	**0.5012**	0.005	1.0	ug/L	11	Standard
	Ag	107	101.3	22.9	**0.0020**	0.003	151.7	ug/L	68	Standard
	Cd	111	246.3	7.9	**0.1051**	0.008	7.2	mg/L	2	Standard
	Cd	114	605.6	11.0	**0.1111**	0.009	8.4	ug/L	12	Standard
>	**In**	115	612194.6	4.4				ug/L	587441	Standard
	Sn	118	1470.1	15.0	**0.1461**	0.027	18.7	ug/L	497	Standard
	Sb	123	727.1	5.4	**0.1452**	0.003	2.2	ug/L	78	Standard
	Ba	135	24533.2	4.5	**11.4855**	0.055	0.5	ug/L	24	Standard
	Ce	140	44839.6	2.7				ug/L	25	Standard
>	**Tb**	159	1002318.5	3.7				ug/L	939701	Standard
	Ho	165	778.4	7.8				ug/L	10	Standard
	Tl	203	123.7	14.4	**0.0129**	0.002	13.5	ug/L	4	Standard
	Tl	205	116.7	9.9	**0.0161**	0.001	9.2	ug/L	2	Standard
	Pb	206	819.0	5.9	**0.0785**	0.008	10.4	ug/L	240	Standard
	Pb	207	658.0	2.8	**0.0655**	0.007	10.9	ug/L	198	Standard
	Pb	208	2733.4	3.4	**0.0682**	0.002	2.2	ug/L	883	Standard
	U	238	29463.0	4.0	**4.1728**	0.071	1.7	ug/L	3	Standard
>	**Bi**	209	536825.9	2.9				ug/L	523984	Standard

Sample ID: L1606098603
Report Date/Time: Wednesday, June 29, 2016 13:15:19
Page 1

Approved: July 04, 2016

Na	23	18.3	31.5	**216.6166**	73.255	33.8	mg/L	0	Standard
Mg	24	7246.7	2.5	**9.4794**	0.084	0.9	mg/L	37	Standard
K	39	393.3	11.0	**0.7727**	0.091	11.8	mg/L	38	Standard
Ca	43	1938.5	3.9	**122.4130**	4.002	3.3	mg/L	10	Standard
Fe	54	1213.7	6.6	**0.9123**	0.078	8.6	mg/L	179	Standard
Fe	57	1360.1	3.5	**3.6434**	0.167	4.6	mg/L	335	Standard
> **Sc-1**	45	59811.2	2.1				mg/L	61029	Standard
Cl	35	41831.0	3.4				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	8062.1	5.5				ug/L	1547	Standard
P	31	19340.7	4.7				ug/L	15286	Standard
S	34	2026.8	7.6				ug/L	1572	Standard
Sr	88	1283.4	13.4				ug/L	73	Standard
C	12	153.3	19.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	1258.4	10.3				mg/L	6	Standard
Ho-1	165	778.4	7.8				mg/L	10	Standard
Er	166	800.0	9.9				mg/L	13	Standard
I	127	76815.4	9.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		143.190	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.817	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098603
Report Date/Time: Wednesday, June 29, 2016 13:15:19
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	104.214
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.451
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098603
Report Date/Time: Wednesday, June 29, 2016 13:15:19
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603PS WG574572-01
Sample Date/Time: Wednesday, June 29, 2016 13:16:14
Number of Replicates: 3
Autosampler Position: 316
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	107768.6	1.4				ug/L	74699	Standard
	Be	9	63154.7	2.1	**49.7293**	0.735	1.5	ug/L	8	Standard
	Al	27	62883924.9	0.4	**296.6156**	4.197	1.4	ug/L	1493	Standard
	Sc	45	60669.6	1.7				ug/L	61029	Standard
	Ti	47	431.3	11.1	**1.0313**	0.120	11.7	ug/L	35	Standard
	V	51	423481.5	0.7	**56.0150**	0.782	1.4	ug/L	1435	Standard
	Cr	52	394353.4	1.0	**52.3183**	0.613	1.2	ug/L	7635	Standard
	Cr	53	48385.5	1.4	**52.1545**	1.086	2.1	ug/L	363	Standard
	Mn	55	12676353.0	0.9	**1825.0939**	28.984	1.6	ug/L	747	Standard
	Co	59	406698.1	0.5	**61.7025**	0.702	1.1	ug/L	234	Standard
	Ni	60	105609.8	0.7	**56.4229**	0.793	1.4	ug/L	215	Standard
	Cu	65	109624.8	0.4	**58.9403**	0.355	0.6	ug/L	104	Standard
	Zn	66	80974.1	1.0	**73.8592**	1.163	1.6	ug/L	363	Standard
>	Ge	72	493944.8	0.7				ug/L	495142	Standard
	As	75	62963.2	0.4	**56.4099**	0.584	1.0	ug/L	-50	Standard
	Se	82	6528.2	1.4	**59.4919**	1.250	2.1	ug/L	14	Standard
	Se-1	77	4208.9	2.0	**57.1456**	1.224	2.1	ug/L	42	Standard
>	Ga	71	341.7	3.7				mg/L	42	Standard
	Rb	85	9437.9	2.3				ug/L	17	Standard
	Y	89	453637.5	2.0				ug/L	417098	Standard
>	Rh	103	333.3	4.8				ug/L	13	Standard
	Mo	98	1406.8	1.9	**0.5023**	0.008	1.6	ug/L	11	Standard
	Ag	107	336751.7	1.1	**49.7241**	0.838	1.7	ug/L	68	Standard
	Cd	111	110645.7	0.7	**50.9305**	0.669	1.3	mg/L	2	Standard
	Cd	114	268194.0	0.5	**50.3333**	0.438	0.9	ug/L	12	Standard
>	In	115	610323.2	0.7				ug/L	587441	Standard
	Sn	118	1538.4	2.3	**0.1588**	0.006	3.5	ug/L	497	Standard
	Sb	123	241381.6	1.1	**49.7617**	0.807	1.6	ug/L	78	Standard
	Ba	135	126394.4	1.4	**59.4274**	1.193	2.0	ug/L	24	Standard
	Ce	140	44334.8	0.8				ug/L	25	Standard
>	Tb	159	999714.8	1.0				ug/L	939701	Standard
	Ho	165	786.7	10.3				ug/L	10	Standard
	Tl	203	445077.8	1.6	**52.3048**	0.918	1.8	ug/L	4	Standard
	Tl	205	409895.5	2.2	**55.5017**	1.265	2.3	ug/L	2	Standard
	Pb	206	293321.0	0.4	**53.1218**	0.228	0.4	ug/L	240	Standard
	Pb	207	252054.7	0.9	**51.3779**	0.617	1.2	ug/L	198	Standard
	Pb	208	1026149.4	0.5	**52.9396**	0.195	0.4	ug/L	883	Standard
	U	238	420484.7	1.6	**59.1737**	0.984	1.7	ug/L	3	Standard
>	Bi	209	540895.8	0.3				ug/L	523984	Standard

Sample ID: L1606098603PS WG574572-01
Report Date/Time: Wednesday, June 29, 2016 13:18:31
Page 1

Approved: July 04, 2016

Na	23	16.7	45.8	**192.4294**	85.849	44.6	mg/L	0	Standard
Mg	24	7256.7	2.5	**9.3609**	0.349	3.7	mg/L	37	Standard
K	39	406.7	8.6	**0.7894**	0.064	8.1	mg/L	38	Standard
Ca	43	1853.4	4.6	**115.2974**	6.640	5.8	mg/L	10	Standard
Fe	54	1091.3	5.4	**0.7921**	0.036	4.5	mg/L	179	Standard
Fe	57	1315.1	6.5	**3.4198**	0.224	6.6	mg/L	335	Standard
> **Sc-1**	45	60669.6	1.7				mg/L	61029	Standard
Cl	35	43150.7	2.4				ug/L	20195	Standard
Kr	83	2.3	49.5				ug/L	4	Standard
Br	81	7932.1	5.6				ug/L	1547	Standard
P	31	18981.9	3.1				ug/L	15286	Standard
S	34	1910.1	1.6				ug/L	1572	Standard
Sr	88	1396.7	9.1				ug/L	73	Standard
C	12	166.7	22.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	1221.7	7.2				mg/L	6	Standard
Ho-1	165	786.7	10.3				mg/L	10	Standard
Er	166	800.0	7.6				mg/L	13	Standard
I	127	76722.5	5.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		144.271	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.758	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098603PS WG574572-01
Report Date/Time: Wednesday, June 29, 2016 13:18:31
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	103.895
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	103.227
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098603PS WG574572-01
Report Date/Time: Wednesday, June 29, 2016 13:18:31
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603SDL WG574572-02
Sample Date/Time: Wednesday, June 29, 2016 13:19:25
Number of Replicates: 3
Autosampler Position: 317
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	109627.0	2.2				ug/L	74699	Standard
	Be	9	55.0	32.8	**0.0501**	0.014	27.0	ug/L	8	Standard
	Al	27	13052757.9	2.1	**60.5215**	1.604	2.7	ug/L	1493	Standard
	Sc	45	62358.1	1.0				ug/L	61029	Standard
	Ti	47	123.0	14.7	**0.1926**	0.050	25.7	ug/L	35	Standard
	V	51	2451.6	1.0	**0.1057**	0.008	7.9	ug/L	1435	Standard
	Cr	52	10290.8	1.8	**0.2648**	0.008	2.8	ug/L	7635	Standard
	Cr	53	588.3	10.7	**0.2610**	0.063	24.3	ug/L	363	Standard
	Mn	55	2629853.6	1.6	**350.4744**	10.232	2.9	ug/L	747	Standard
	Co	59	4233.3	3.1	**0.5577**	0.018	3.3	ug/L	234	Standard
	Ni	60	2886.6	1.8	**1.3071**	0.037	2.8	ug/L	215	Standard
	Cu	65	3859.5	1.6	**1.8207**	0.021	1.2	ug/L	104	Standard
	Zn	66	7087.3	1.2	**5.5680**	0.153	2.7	ug/L	363	Standard
>	**Ge**	72	533577.2	1.8				ug/L	495142	Standard
	As	75	314.5	3.2	**0.3203**	0.010	3.2	ug/L	-50	Standard
	Se	82	28.2	16.0	**0.1004**	0.037	37.2	ug/L	14	Standard
	Se-1	77	51.7	9.9	**0.0349**	0.075	214.8	ug/L	42	Standard
>	**Ga**	71	153.3	21.0				mg/L	42	Standard
	Rb	85	1916.8	2.5				ug/L	17	Standard
	Y	89	457911.1	1.7				ug/L	417098	Standard
>	**Rh**	103	73.3	17.2				ug/L	13	Standard
	Mo	98	334.8	7.8	**0.1120**	0.009	8.5	ug/L	11	Standard
	Ag	107	142.0	14.1	**0.0075**	0.003	37.0	ug/L	68	Standard
	Cd	111	78.8	16.1	**0.0274**	0.006	20.4	mg/L	2	Standard
	Cd	114	171.7	29.4	**0.0293**	0.009	31.0	ug/L	12	Standard
>	**In**	115	626915.6	0.6				ug/L	587441	Standard
	Sn	118	618.3	15.8	**0.0044**	0.015	338.7	ug/L	497	Standard
	Sb	123	2031.6	21.7	**0.4032**	0.086	21.3	ug/L	78	Standard
	Ba	135	5054.8	0.4	**2.2983**	0.024	1.0	ug/L	24	Standard
	Ce	140	9302.9	0.9				ug/L	25	Standard
>	**Tb**	159	1014481.9	2.3				ug/L	939701	Standard
	Ho	165	175.0	2.9				ug/L	10	Standard
	Tl	203	125.0	17.4	**0.0120**	0.002	18.4	ug/L	4	Standard
	Tl	205	85.0	15.6	**0.0110**	0.002	15.7	ug/L	2	Standard
	Pb	206	501.3	9.2	**0.0141**	0.008	53.2	ug/L	240	Standard
	Pb	207	392.0	5.3	**0.0051**	0.005	98.5	ug/L	198	Standard
	Pb	208	1534.0	5.0	**0.0003**	0.005	1457.8	ug/L	883	Standard
	U	238	6033.2	1.4	**0.7925**	0.020	2.5	ug/L	3	Standard
>	**Bi**	209	576425.9	1.5				ug/L	523984	Standard

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:21:42
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**37.9583**	65.737	173.2	mg/L	0	Standard
Mg	24	1580.1	4.3	**1.9318**	0.105	5.4	mg/L	37	Standard
K	39	138.3	24.1	**0.1853**	0.073	39.3	mg/L	38	Standard
Ca	43	413.3	5.7	**23.0292**	1.446	6.3	mg/L	10	Standard
Fe	54	367.9	5.4	**0.1671**	0.019	11.4	mg/L	179	Standard
Fe	57	475.0	7.9	**0.5062**	0.132	26.1	mg/L	335	Standard
Sc-1	45	62358.1	1.0				mg/L	61029	Standard
Cl	35	34619.5	1.5				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	3197.0	3.6				ug/L	1547	Standard
P	31	18749.9	1.5				ug/L	15286	Standard
S	34	2010.1	3.0				ug/L	1572	Standard
Sr	88	305.0	20.5				ug/L	73	Standard
C	12	113.3	31.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	348.8	17.3				mg/L	6	Standard
Ho-1	165	175.0	2.9				mg/L	10	Standard
Er	166	163.3	52.8				mg/L	13	Standard
I	127	19115.4	2.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		146.759	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		107.762	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:21:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.720
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.008
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603SDL WG574572-02
Sample Date/Time: Wednesday, June 29, 2016 13:22:37
Number of Replicates: 3
Autosampler Position: 318
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	107882.7	0.8				ug/L	74699	Standard
	Be	9	30.0	44.1	**0.0312**	0.010	32.7	ug/L	8	Standard
	Al	27	2675117.5	0.8	**12.5917**	0.213	1.7	ug/L	1493	Standard
	Sc	45	65760.7	1.1				ug/L	61029	Standard
	Ti	47	48.3	10.6	**0.0072**	0.012	160.2	ug/L	35	Standard
	V	51	2266.6	5.1	**0.0781**	0.015	19.5	ug/L	1435	Standard
	Cr	52	10143.4	2.6	**0.2243**	0.020	8.9	ug/L	7635	Standard
	Cr	53	413.3	17.1	**0.0775**	0.065	83.8	ug/L	363	Standard
	Mn	55	539662.1	1.4	**70.5366**	0.569	0.8	ug/L	747	Standard
	Co	59	1015.0	2.6	**0.1030**	0.006	5.9	ug/L	234	Standard
	Ni	60	769.4	2.6	**0.2509**	0.010	4.1	ug/L	215	Standard
	Cu	65	948.0	4.2	**0.3607**	0.016	4.4	ug/L	104	Standard
	Zn	66	2642.9	1.7	**1.7518**	0.018	1.1	ug/L	363	Standard
>	**Ge**	72	543000.4	1.8				ug/L	495142	Standard
	As	75	12.5	225.4	**0.0698**	0.023	32.5	ug/L	-50	Standard
	Se	82	15.8	49.7	**-0.0063**	0.066	1051.8	ug/L	14	Standard
	Se-1	77	45.3	17.1	**-0.0563**	0.096	170.3	ug/L	42	Standard
>	**Ga**	71	85.0	15.6				mg/L	42	Standard
	Rb	85	428.3	11.2				ug/L	17	Standard
	Y	89	460200.9	1.8				ug/L	417098	Standard
>	**Rh**	103	35.0	24.7				ug/L	13	Standard
	Mo	98	83.6	11.1	**0.0226**	0.003	14.3	ug/L	11	Standard
	Ag	107	112.7	14.2	**0.0027**	0.002	80.0	ug/L	68	Standard
	Cd	111	20.9	22.0	**0.0011**	0.002	176.9	mg/L	2	Standard
	Cd	114	53.5	39.5	**0.0073**	0.004	50.6	ug/L	12	Standard
>	**In**	115	649578.1	0.2				ug/L	587441	Standard
	Sn	118	553.3	9.5	**-0.0090**	0.008	92.4	ug/L	497	Standard
	Sb	123	558.9	31.9	**0.1041**	0.035	33.4	ug/L	78	Standard
	Ba	135	1064.4	0.7	**0.4543**	0.003	0.7	ug/L	24	Standard
	Ce	140	1928.5	4.3				ug/L	25	Standard
>	**Tb**	159	1049822.6	1.3				ug/L	939701	Standard
	Ho	165	40.0	33.1				ug/L	10	Standard
	Tl	203	40.7	35.8	**0.0026**	0.002	59.7	ug/L	4	Standard
	Tl	205	45.0	38.5	**0.0057**	0.002	36.9	ug/L	2	Standard
	Pb	206	372.0	6.6	**-0.0102**	0.004	38.7	ug/L	240	Standard
	Pb	207	294.7	6.9	**-0.0156**	0.004	25.6	ug/L	198	Standard
	Pb	208	1159.4	2.3	**-0.0199**	0.001	7.3	ug/L	883	Standard
	U	238	1206.0	0.9	**0.1490**	0.002	1.2	ug/L	3	Standard
>	**Bi**	209	598311.7	0.5				ug/L	523984	Standard

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:24:54
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.6400**	30.545	173.2	mg/L	0	Standard
Mg	24	363.3	29.0	**0.3711**	0.131	35.3	mg/L	37	Standard
K	39	63.3	27.7	**0.0157**	0.035	226.7	mg/L	38	Standard
Ca	43	110.0	12.0	**3.9202**	0.707	18.0	mg/L	10	Standard
Fe	54	190.5	10.6	**0.0117**	0.017	142.3	mg/L	179	Standard
Fe	57	268.3	12.0	**-0.2271**	0.110	48.6	mg/L	335	Standard
> **Sc-1**	45	65760.7	1.1				mg/L	61029	Standard
Cl	35	34122.4	1.0				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	2286.8	8.5				ug/L	1547	Standard
P	31	19529.2	1.3				ug/L	15286	Standard
S	34	2301.8	4.6				ug/L	1572	Standard
Sr	88	130.0	32.9				ug/L	73	Standard
C	12	83.3	6.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	60.8	35.4				mg/L	6	Standard
Ho-1	165	40.0	33.1				mg/L	10	Standard
Er	166	53.3	28.6				mg/L	13	Standard
I	127	5874.5	8.0				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		144.424	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		109.666	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:24:54
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.578
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		114.185
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 13:25:50
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	96295.4	4.3				ug/L	74699	Standard
	Be	9	56739.5	4.7	**50.0144**	1.617	3.2	ug/L	8	Standard
	Al	27	8198811.7	1.9	**43.2979**	1.271	2.9	ug/L	1493	Standard
	Sc	45	60020.3	2.0				ug/L	61029	Standard
	Ti	47	38068.4	0.4	**97.0435**	0.523	0.5	ug/L	35	Standard
	V	51	407671.2	1.0	**52.0137**	0.145	0.3	ug/L	1435	Standard
	Cr	52	382711.8	0.9	**48.9253**	0.135	0.3	ug/L	7635	Standard
	Cr	53	47653.2	2.5	**49.5394**	1.027	2.1	ug/L	363	Standard
	Mn	55	366963.7	3.6	**50.8236**	1.362	2.7	ug/L	747	Standard
	Co	59	380449.1	2.9	**55.6813**	1.097	2.0	ug/L	234	Standard
	Ni	60	94585.7	1.1	**48.7401**	0.435	0.9	ug/L	215	Standard
	Cu	65	94763.8	1.5	**49.1432**	0.553	1.1	ug/L	104	Standard
	Zn	66	56752.8	2.3	**49.7949**	0.707	1.4	ug/L	363	Standard
>	Ge	72	511914.0	0.9				ug/L	495142	Standard
	As	75	57985.2	1.9	**50.1285**	0.568	1.1	ug/L	-50	Standard
	Se	82	5989.4	2.5	**52.6418**	0.964	1.8	ug/L	14	Standard
	Se-1	77	3921.2	2.5	**51.3010**	0.802	1.6	ug/L	42	Standard
>	Ga	71	220.0	19.4				mg/L	42	Standard
	Rb	85	765.0	2.0				ug/L	17	Standard
	Y	89	446724.4	3.3				ug/L	417098	Standard
>	Rh	103	20.0	43.3				ug/L	13	Standard
	Mo	98	280984.3	2.1	**102.0498**	2.098	2.1	ug/L	11	Standard
	Ag	107	335763.8	0.9	**49.8677**	1.010	2.0	ug/L	68	Standard
	Cd	111	105816.2	1.8	**48.9835**	0.569	1.2	mg/L	2	Standard
	Cd	114	257781.1	1.6	**48.6563**	0.589	1.2	ug/L	12	Standard
>	In	115	606960.1	2.8				ug/L	587441	Standard
	Sn	118	295693.2	2.0	**48.9061**	0.749	1.5	ug/L	497	Standard
	Sb	123	236523.1	0.9	**49.0438**	0.901	1.8	ug/L	78	Standard
	Ba	135	99992.2	1.5	**47.2781**	0.609	1.3	ug/L	24	Standard
	Ce	140	91.7	6.3				ug/L	25	Standard
>	Tb	159	985978.7	1.6				ug/L	939701	Standard
	Ho	165	11.7	89.2				ug/L	10	Standard
	Tl	203	433708.5	1.0	**49.3829**	0.257	0.5	ug/L	4	Standard
	Tl	205	391770.8	1.1	**51.3967**	0.163	0.3	ug/L	2	Standard
	Pb	206	284957.7	0.7	**50.0002**	0.477	1.0	ug/L	240	Standard
	Pb	207	255175.6	1.7	**50.3910**	0.211	0.4	ug/L	198	Standard
	Pb	208	1010912.5	1.9	**50.5251**	0.466	0.9	ug/L	883	Standard
	U	238	378758.8	0.6	**51.6451**	0.416	0.8	ug/L	3	Standard
>	Bi	209	558275.4	1.4				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 13:28:07
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4168.9	2.3	**5.4065**	0.021	0.4	mg/L	37	Standard
K	39	1983.5	2.5	**4.3432**	0.044	1.0	mg/L	38	Standard
Ca	43	106.7	26.7	**4.3258**	1.829	42.3	mg/L	10	Standard
Fe	54	6165.0	5.0	**5.1771**	0.265	5.1	mg/L	179	Standard
Fe	57	1720.1	9.9	**4.8644**	0.466	9.6	mg/L	335	Standard
> **Sc-1**	45	60020.3	2.0				mg/L	61029	Standard
Cl	35	28213.0	5.7				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	1853.5	23.0				ug/L	1547	Standard
P	31	20223.5	3.5				ug/L	15286	Standard
S	34	2575.2	6.1				ug/L	1572	Standard
Sr	88	78.3	26.6				ug/L	73	Standard
C	12	100.0	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	9.2	116.5				mg/L	6	Standard
Ho-1	165	11.7	89.2				mg/L	10	Standard
Er	166	16.7	91.7				mg/L	13	Standard
I	127	2436.9	25.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	100.029		
Al	27	86.596		
Sc	45			
Ti	47	97.044		
V	51	104.027		
Cr	52	97.851		
Cr	53			
Mn	55	101.647		
Co	59	111.363		
Ni	60	97.480		
Cu	65	98.286		
Zn	66	99.590		
> Ge	72		103.387	
As	75	100.257		
Se	82	105.284		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 13:28:07
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.050		
Ag	107	99.735		
Cd	111	97.967		
Cd	114			
In	115		103.323	
Sn	118	97.812		
Sb	123	98.088		
Ba	135	94.556		
Ce	140			
Tb	159			
Ho	165			
Tl	203	98.766		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.050		
U	238	103.290		
Bi	209		106.544	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 13:28:07
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 13:29:01
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94691.7	1.8				ug/L	74699	Standard
	Be	9	76.7	106.2	**0.0755**	0.071	94.2	ug/L	8	Standard
	Al	27	36843.1	136.1	**0.1814**	0.263	145.0	ug/L	1493	Standard
	Sc	45	60632.8	2.3				ug/L	61029	Standard
	Ti	47	43.0	36.1	**-0.0010**	0.038	3637.8	ug/L	35	Standard
	V	51	1519.7	11.1	**-0.0035**	0.019	532.9	ug/L	1435	Standard
	Cr	52	7045.6	5.2	**-0.1157**	0.044	37.8	ug/L	7635	Standard
	Cr	53	391.7	11.6	**0.0747**	0.044	59.2	ug/L	363	Standard
	Mn	55	7226.9	125.7	**0.8342**	1.228	147.2	ug/L	747	Standard
	Co	59	517.7	90.4	**0.0371**	0.066	179.1	ug/L	234	Standard
	Ni	60	246.3	39.3	**0.0014**	0.048	3317.1	ug/L	215	Standard
	Cu	65	191.7	56.6	**-0.0057**	0.054	949.5	ug/L	104	Standard
	Zn	66	417.3	10.8	**-0.0898**	0.036	40.1	ug/L	363	Standard
>	**Ge**	72	518398.0	1.4				ug/L	495142	Standard
	As	75	-30.5	76.0	**0.0334**	0.020	59.9	ug/L	-50	Standard
	Se	82	13.4	12.4	**-0.0212**	0.016	76.0	ug/L	14	Standard
	Se-1	77	53.3	15.9	**0.0746**	0.106	141.7	ug/L	42	Standard
>	**Ga**	71	155.0	18.0				mg/L	42	Standard
	Rb	85	21.7	93.3				ug/L	17	Standard
	Y	89	436745.4	0.7				ug/L	417098	Standard
>	**Rh**	103	15.0	33.3				ug/L	13	Standard
	Mo	98	315.2	32.1	**0.1069**	0.036	33.6	ug/L	11	Standard
	Ag	107	222.7	75.6	**0.0196**	0.025	125.0	ug/L	68	Standard
	Cd	111	57.8	114.5	**0.0184**	0.030	163.6	mg/L	2	Standard
	Cd	114	159.0	130.1	**0.0274**	0.038	140.1	ug/L	12	Standard
>	**In**	115	616374.3	0.2				ug/L	587441	Standard
	Sn	118	845.0	17.1	**0.0432**	0.024	55.2	ug/L	497	Standard
	Sb	123	593.6	13.2	**0.1170**	0.016	13.9	ug/L	78	Standard
	Ba	135	102.7	102.4	**0.0317**	0.049	153.8	ug/L	24	Standard
	Ce	140	58.3	55.8				ug/L	25	Standard
>	**Tb**	159	995160.6	1.6				ug/L	939701	Standard
	Ho	165	15.0	100.0				ug/L	10	Standard
	Tl	203	219.0	114.2	**0.0226**	0.028	122.3	ug/L	4	Standard
	Tl	205	381.7	141.6	**0.0491**	0.069	140.6	ug/L	2	Standard
	Pb	206	388.0	38.6	**-0.0042**	0.025	597.1	ug/L	240	Standard
	Pb	207	336.0	35.9	**-0.0047**	0.023	485.4	ug/L	198	Standard
	Pb	208	2012.1	79.6	**0.0246**	0.078	315.7	ug/L	883	Standard
	U	238	395.3	145.1	**0.0482**	0.076	157.9	ug/L	3	Standard
>	**Bi**	209	567355.9	1.2				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 13:31:18
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	58.3	34.6	**0.0111**	0.028	248.5	mg/L	37	Standard
K	39	40.0	12.5	**-0.0251**	0.012	46.0	mg/L	38	Standard
Ca	43	31.7	9.1	**-0.5078**	0.213	41.9	mg/L	10	Standard
Fe	54	155.5	6.6	**-0.0056**	0.009	155.5	mg/L	179	Standard
Fe	57	230.0	7.8	**-0.2862**	0.078	27.3	mg/L	335	Standard
> **Sc-1**	45	60632.8	2.3				mg/L	61029	Standard
Cl	35	27178.4	0.6				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	1736.8	2.2				ug/L	1547	Standard
P	31	19502.5	1.8				ug/L	15286	Standard
S	34	2301.8	3.7				ug/L	1572	Standard
Sr	88	75.0	26.7				ug/L	73	Standard
C	12	86.7	17.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	12.7	93.2				mg/L	6	Standard
Ho-1	165	15.0	100.0				mg/L	10	Standard
Er	166	13.3	86.6				mg/L	13	Standard
I	127	1935.1	25.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		104.697	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 13:31:18
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.925
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.277
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Mn	55	

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 13:31:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098605
Sample Date/Time: Wednesday, June 29, 2016 13:32:14
Number of Replicates: 3
Autosampler Position: 319
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93251.1	1.7				ug/L	74699	Standard
	Be	9	326.7	10.4	**0.3052**	0.035	11.6	ug/L	8	Standard
	Al	27	19740200.2	0.7	**107.5996**	1.185	1.1	ug/L	1493	Standard
	Sc	45	60365.1	2.3				ug/L	61029	Standard
	Ti	47	450.3	3.0	**1.0603**	0.030	2.8	ug/L	35	Standard
	V	51	4644.9	8.0	**0.4097**	0.047	11.5	ug/L	1435	Standard
	Cr	52	12104.2	0.3	**0.5838**	0.010	1.7	ug/L	7635	Standard
	Cr	53	928.4	3.1	**0.6586**	0.026	3.9	ug/L	363	Standard
	Mn	55	98559.3	4.8	**13.7841**	0.650	4.7	ug/L	747	Standard
	Co	59	3296.0	4.6	**0.4539**	0.021	4.7	ug/L	234	Standard
	Ni	60	3163.0	1.8	**1.5389**	0.038	2.4	ug/L	215	Standard
	Cu	65	32310.7	0.8	**16.9827**	0.169	1.0	ug/L	104	Standard
	Zn	66	47105.4	0.5	**41.9893**	0.365	0.9	ug/L	363	Standard
>	**Ge**	72	503086.3	0.5				ug/L	495142	Standard
	As	75	294.8	15.7	**0.3186**	0.040	12.5	ug/L	-50	Standard
	Se	82	52.3	21.7	**0.3309**	0.101	30.6	ug/L	14	Standard
	Se-1	77	70.7	11.4	**0.3303**	0.114	34.4	ug/L	42	Standard
>	**Ga**	71	295.0	9.0				mg/L	42	Standard
	Rb	85	9109.4	3.7				ug/L	17	Standard
	Y	89	448266.6	2.2				ug/L	417098	Standard
>	**Rh**	103	83.3	28.4				ug/L	13	Standard
	Mo	98	1843.0	5.2	**0.6500**	0.034	5.3	ug/L	11	Standard
	Ag	107	226.3	74.5	**0.0200**	0.024	121.9	ug/L	68	Standard
	Cd	111	104.1	84.7	**0.0392**	0.040	101.3	mg/L	2	Standard
	Cd	114	254.7	74.3	**0.0449**	0.035	77.3	ug/L	12	Standard
>	**In**	115	619535.9	1.6				ug/L	587441	Standard
	Sn	118	1750.1	13.2	**0.1895**	0.038	20.1	ug/L	497	Standard
	Sb	123	790.9	17.9	**0.1564**	0.028	18.0	ug/L	78	Standard
	Ba	135	72683.9	1.3	**33.6664**	0.969	2.9	ug/L	24	Standard
	Ce	140	5139.2	3.4				ug/L	25	Standard
>	**Tb**	159	985815.7	2.2				ug/L	939701	Standard
	Ho	165	223.3	9.3				ug/L	10	Standard
	Tl	203	317.0	96.7	**0.0353**	0.035	100.6	ug/L	4	Standard
	Tl	205	193.3	47.9	**0.0263**	0.012	46.4	ug/L	2	Standard
	Pb	206	1382.4	8.3	**0.1795**	0.018	10.0	ug/L	240	Standard
	Pb	207	1152.7	10.0	**0.1653**	0.022	13.1	ug/L	198	Standard
	Pb	208	4722.3	7.7	**0.1699**	0.018	10.3	ug/L	883	Standard
	U	238	9861.9	3.5	**1.3841**	0.043	3.1	ug/L	3	Standard
>	**Bi**	209	540670.7	1.7				ug/L	523984	Standard

Sample ID: L1606098605
Report Date/Time: Wednesday, June 29, 2016 13:34:31
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	6986.6	1.5	**9.0573**	0.310	3.4	mg/L	37	Standard
K	39	308.3	17.9	**0.5733**	0.109	19.0	mg/L	38	Standard
Ca	43	606.7	4.1	**36.2577**	2.485	6.9	mg/L	10	Standard
Fe	54	348.6	15.4	**0.1605**	0.045	28.3	mg/L	179	Standard
Fe	57	585.0	8.4	**0.9378**	0.194	20.7	mg/L	335	Standard
> **Sc-1**	45	60365.1	2.3				mg/L	61029	Standard
Cl	35	31561.5	3.4				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	4964.2	11.3				ug/L	1547	Standard
P	31	20053.3	1.8				ug/L	15286	Standard
S	34	2155.2	4.8				ug/L	1572	Standard
Sr	88	343.3	18.3				ug/L	73	Standard
C	12	163.3	34.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	294.3	16.7				mg/L	6	Standard
Ho-1	165	223.3	9.3				mg/L	10	Standard
Er	166	120.0	16.7				mg/L	13	Standard
I	127	9371.2	3.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		124.836	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.604	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098605
Report Date/Time: Wednesday, June 29, 2016 13:34:31
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		105.463
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.184
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098607
Sample Date/Time: Wednesday, June 29, 2016 13:35:26
Number of Replicates: 3
Autosampler Position: 320
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96566.8	1.9				ug/L	74699	Standard
	Be	9	85.0	29.4	**0.0821**	0.021	25.7	ug/L	8	Standard
	Al	27	20333697.9	3.0	**107.0230**	2.931	2.7	ug/L	1493	Standard
	Sc	45	62230.9	2.2				ug/L	61029	Standard
	Ti	47	287.7	7.2	**0.6119**	0.056	9.2	ug/L	35	Standard
	V	51	3941.3	2.6	**0.3001**	0.017	5.7	ug/L	1435	Standard
	Cr	52	11512.0	2.6	**0.4518**	0.053	11.8	ug/L	7635	Standard
	Cr	53	838.4	4.6	**0.5317**	0.044	8.4	ug/L	363	Standard
	Mn	55	502202.8	0.8	**68.3542**	0.721	1.1	ug/L	747	Standard
	Co	59	4570.7	1.4	**0.6201**	0.010	1.6	ug/L	234	Standard
	Ni	60	3293.0	0.7	**1.5464**	0.014	0.9	ug/L	215	Standard
	Cu	65	6694.1	1.0	**3.3123**	0.070	2.1	ug/L	104	Standard
	Zn	66	11579.4	0.2	**9.6129**	0.092	1.0	ug/L	363	Standard
>	**Ge**	72	521399.1	1.0				ug/L	495142	Standard
	As	75	65.2	31.2	**0.1148**	0.017	14.9	ug/L	-50	Standard
	Se	82	28.9	16.0	**0.1117**	0.041	36.5	ug/L	14	Standard
	Se-1	77	63.7	8.1	**0.2056**	0.068	33.2	ug/L	42	Standard
>	**Ga**	71	136.7	35.5				mg/L	42	Standard
	Rb	85	5054.2	2.4				ug/L	17	Standard
	Y	89	452894.0	2.8				ug/L	417098	Standard
>	**Rh**	103	165.0	28.9				ug/L	13	Standard
	Mo	98	1211.0	1.7	**0.4303**	0.008	1.9	ug/L	11	Standard
	Ag	107	106.3	11.0	**0.0028**	0.002	64.6	ug/L	68	Standard
	Cd	111	72.3	5.5	**0.0253**	0.002	7.2	mg/L	2	Standard
	Cd	114	189.4	18.5	**0.0333**	0.006	19.3	ug/L	12	Standard
>	**In**	115	612132.4	0.4				ug/L	587441	Standard
	Sn	118	1080.0	6.7	**0.0827**	0.012	15.0	ug/L	497	Standard
	Sb	123	567.1	13.2	**0.1124**	0.015	13.5	ug/L	78	Standard
	Ba	135	35457.1	0.1	**16.6089**	0.050	0.3	ug/L	24	Standard
	Ce	140	1880.1	4.7				ug/L	25	Standard
>	**Tb**	159	996852.4	0.7				ug/L	939701	Standard
	Ho	165	58.3	17.8				ug/L	10	Standard
	Tl	203	60.0	16.9	**0.0051**	0.001	23.4	ug/L	4	Standard
	Tl	205	60.0	16.7	**0.0081**	0.001	15.8	ug/L	2	Standard
	Pb	206	636.0	0.7	**0.0415**	0.001	3.4	ug/L	240	Standard
	Pb	207	529.0	3.2	**0.0355**	0.003	9.7	ug/L	198	Standard
	Pb	208	2291.4	2.0	**0.0416**	0.003	7.7	ug/L	883	Standard
	U	238	20180.8	2.1	**2.7702**	0.079	2.9	ug/L	3	Standard
>	**Bi**	209	553749.9	0.8				ug/L	523984	Standard

Sample ID: L1606098607
Report Date/Time: Wednesday, June 29, 2016 13:37:42
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**57.0696**	57.922	101.5	mg/L	0	Standard
Mg	24	5551.0	3.6	**6.9630**	0.244	3.5	mg/L	37	Standard
K	39	316.7	11.2	**0.5727**	0.081	14.1	mg/L	38	Standard
Ca	43	636.7	7.9	**36.9594**	3.698	10.0	mg/L	10	Standard
Fe	54	176.9	14.5	**0.0085**	0.018	212.8	mg/L	179	Standard
Fe	57	531.7	3.6	**0.6989**	0.084	12.0	mg/L	335	Standard
> **Sc-1**	45	62230.9	2.2				mg/L	61029	Standard
Cl	35	31386.5	1.7				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	3093.6	2.6				ug/L	1547	Standard
P	31	20912.8	5.9				ug/L	15286	Standard
S	34	2228.5	4.5				ug/L	1572	Standard
Sr	88	703.3	5.1				ug/L	73	Standard
C	12	126.7	40.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	102.2	30.1				mg/L	6	Standard
Ho-1	165	58.3	17.8				mg/L	10	Standard
Er	166	93.3	6.2				mg/L	13	Standard
I	127	29892.9	5.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		129.275	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		105.303	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098607
Report Date/Time: Wednesday, June 29, 2016 13:37:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.203
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.681
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098609
Sample Date/Time: Wednesday, June 29, 2016 13:38:37
Number of Replicates: 3
Autosampler Position: 321
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96550.0	0.2				ug/L	74699	Standard
	Be	9	53.3	14.3	**0.0545**	0.007	12.5	ug/L	8	Standard
	Al	27	22878376.4	1.6	**120.4307**	1.945	1.6	ug/L	1493	Standard
	Sc	45	62031.7	0.8				ug/L	61029	Standard
	Ti	47	1850.4	1.6	**4.6226**	0.192	4.2	ug/L	35	Standard
	V	51	5461.8	0.6	**0.5052**	0.022	4.3	ug/L	1435	Standard
	Cr	52	13396.9	1.2	**0.7280**	0.058	8.0	ug/L	7635	Standard
	Cr	53	1881.8	4.3	**1.6444**	0.138	8.4	ug/L	363	Standard
	Mn	55	79733.3	1.7	**10.9451**	0.135	1.2	ug/L	747	Standard
	Co	59	2496.2	2.4	**0.3289**	0.007	2.0	ug/L	234	Standard
	Ni	60	4821.8	0.8	**2.3714**	0.068	2.9	ug/L	215	Standard
	Cu	65	4208.3	2.7	**2.0873**	0.088	4.2	ug/L	104	Standard
	Zn	66	20900.1	0.5	**18.0858**	0.567	3.1	ug/L	363	Standard
>	**Ge**	72	511251.4	2.6				ug/L	495142	Standard
	As	75	201.6	14.3	**0.2338**	0.023	9.7	ug/L	-50	Standard
	Se	82	54.1	5.4	**0.3394**	0.016	4.7	ug/L	14	Standard
	Se-1	77	103.0	11.2	**0.7418**	0.124	16.8	ug/L	42	Standard
>	**Ga**	71	510.0	8.0				mg/L	42	Standard
	Rb	85	13392.6	2.9				ug/L	17	Standard
	Y	89	437578.5	1.9				ug/L	417098	Standard
>	**Rh**	103	115.0	21.7				ug/L	13	Standard
	Mo	98	379.6	2.6	**0.1321**	0.004	2.9	ug/L	11	Standard
	Ag	107	107.7	0.5	**0.0031**	0.000	8.0	ug/L	68	Standard
	Cd	111	160.7	1.3	**0.0664**	0.001	1.0	mg/L	2	Standard
	Cd	114	395.9	2.3	**0.0726**	0.002	3.0	ug/L	12	Standard
>	**In**	115	607119.7	1.0				ug/L	587441	Standard
	Sn	118	1270.1	2.5	**0.1157**	0.007	5.8	ug/L	497	Standard
	Sb	123	941.4	6.3	**0.1909**	0.012	6.2	ug/L	78	Standard
	Ba	135	18081.1	0.3	**8.5321**	0.064	0.8	ug/L	24	Standard
	Ce	140	9564.7	2.3				ug/L	25	Standard
>	**Tb**	159	994932.9	1.3				ug/L	939701	Standard
	Ho	165	368.3	5.1				ug/L	10	Standard
	Tl	203	147.3	4.5	**0.0153**	0.001	4.3	ug/L	4	Standard
	Tl	205	150.0	15.3	**0.0202**	0.003	16.2	ug/L	2	Standard
	Pb	206	1807.1	1.2	**0.2512**	0.001	0.3	ug/L	240	Standard
	Pb	207	1489.4	1.7	**0.2291**	0.005	2.2	ug/L	198	Standard
	Pb	208	6035.1	2.4	**0.2326**	0.004	1.7	ug/L	883	Standard
	U	238	2872.6	2.3	**0.3933**	0.005	1.2	ug/L	3	Standard
>	**Bi**	209	549776.1	1.1				ug/L	523984	Standard

Sample ID: L1606098609
Report Date/Time: Wednesday, June 29, 2016 13:40:55
Page 1

Approved: July 04, 2016

Na	23	6.7	114.6	**75.5001**	86.744	114.9	mg/L	0	Standard
Mg	24	5847.8	3.3	**7.3629**	0.301	4.1	mg/L	37	Standard
K	39	155.0	19.6	**0.2229**	0.066	29.5	mg/L	38	Standard
Ca	43	803.4	3.4	**47.4095**	2.024	4.3	mg/L	10	Standard
Fe	54	459.0	5.6	**0.2447**	0.024	9.9	mg/L	179	Standard
Fe	57	691.7	11.1	**1.2400**	0.273	22.0	mg/L	335	Standard
> **Sc-1**	45	62031.7	0.8				mg/L	61029	Standard
Cl	35	35055.2	1.5				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	8335.6	0.3				ug/L	1547	Standard
P	31	19207.2	4.0				ug/L	15286	Standard
S	34	2091.8	7.3				ug/L	1572	Standard
Sr	88	440.0	15.9				ug/L	73	Standard
C	12	146.7	7.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	580.9	19.4				mg/L	6	Standard
Ho-1	165	368.3	5.1				mg/L	10	Standard
Er	166	330.0	13.9				mg/L	13	Standard
I	127	13649.5	1.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		129.252	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.253	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098609
Report Date/Time: Wednesday, June 29, 2016 13:40:55
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.350
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.922
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098611
Sample Date/Time: Wednesday, June 29, 2016 13:41:50
Number of Replicates: 3
Autosampler Position: 322
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94928.2	1.7				ug/L	74699	Standard
	Be	9	611.7	2.1	**0.5543**	0.002	0.4	ug/L	8	Standard
	Al	27	35531075.2	1.8	**190.3104**	6.477	3.4	ug/L	1493	Standard
	Sc	45	58926.0	1.1				ug/L	61029	Standard
	Ti	47	1221.7	3.3	**3.0767**	0.071	2.3	ug/L	35	Standard
	V	51	7317.6	1.4	**0.7622**	0.017	2.2	ug/L	1435	Standard
	Cr	52	16086.9	1.7	**1.1219**	0.012	1.1	ug/L	7635	Standard
	Cr	53	9996.6	2.0	**10.3612**	0.096	0.9	ug/L	363	Standard
	Mn	55	583301.8	0.5	**82.6724**	0.812	1.0	ug/L	747	Standard
	Co	59	14960.7	0.8	**2.2024**	0.031	1.4	ug/L	234	Standard
	Ni	60	6882.6	4.3	**3.5089**	0.114	3.2	ug/L	215	Standard
	Cu	65	67223.1	1.5	**35.5977**	0.129	0.4	ug/L	104	Standard
	Zn	66	37202.6	1.0	**33.2122**	0.242	0.7	ug/L	363	Standard
>	**Ge**	72	500908.0	1.1				ug/L	495142	Standard
	As	75	1307.8	5.6	**1.2134**	0.054	4.4	ug/L	-50	Standard
	Se	82	362.7	2.0	**3.1282**	0.043	1.4	ug/L	14	Standard
	Se-1	77	555.0	5.1	**6.8873**	0.307	4.5	ug/L	42	Standard
>	**Ga**	71	435.0	11.9				mg/L	42	Standard
	Rb	85	18344.4	2.4				ug/L	17	Standard
	Y	89	436166.7	1.7				ug/L	417098	Standard
>	**Rh**	103	445.0	7.4				ug/L	13	Standard
	Mo	98	3054.9	0.8	**1.1522**	0.023	2.0	ug/L	11	Standard
	Ag	107	121.7	12.3	**0.0060**	0.002	41.3	ug/L	68	Standard
	Cd	111	41.7	14.4	**0.0123**	0.003	24.8	mg/L	2	Standard
	Cd	114	142.6	14.6	**0.0260**	0.004	14.4	ug/L	12	Standard
>	**In**	115	581588.5	1.2				ug/L	587441	Standard
	Sn	118	2201.8	2.3	**0.2861**	0.011	3.9	ug/L	497	Standard
	Sb	123	1347.5	5.7	**0.2874**	0.019	6.6	ug/L	78	Standard
	Ba	135	63997.2	1.5	**31.5730**	0.849	2.7	ug/L	24	Standard
	Ce	140	22939.0	3.1				ug/L	25	Standard
>	**Tb**	159	946929.2	0.6				ug/L	939701	Standard
	Ho	165	348.3	12.2				ug/L	10	Standard
	Tl	203	270.7	7.5	**0.0312**	0.002	6.9	ug/L	4	Standard
	Tl	205	231.7	2.5	**0.0327**	0.001	2.0	ug/L	2	Standard
	Pb	206	3400.0	1.8	**0.5672**	0.008	1.4	ug/L	240	Standard
	Pb	207	2794.6	0.7	**0.5207**	0.008	1.5	ug/L	198	Standard
	Pb	208	11397.0	1.2	**0.5356**	0.008	1.6	ug/L	883	Standard
	U	238	7265.4	1.1	**1.0543**	0.007	0.7	ug/L	3	Standard
>	**Bi**	209	522379.2	0.9				ug/L	523984	Standard

Sample ID: L1606098611
Report Date/Time: Wednesday, June 29, 2016 13:44:07
Page 1

Approved: July 04, 2016

Na	23	10.0	50.0	**119.8009**	60.997	50.9	mg/L	0	Standard
Mg	24	10421.9	2.0	**13.8711**	0.425	3.1	mg/L	37	Standard
K	39	453.3	5.1	**0.9237**	0.053	5.7	mg/L	38	Standard
Ca	43	1046.7	6.1	**65.9862**	4.937	7.5	mg/L	10	Standard
Fe	54	728.6	2.4	**0.5015**	0.016	3.2	mg/L	179	Standard
Fe	57	860.0	12.2	**1.9559**	0.399	20.4	mg/L	335	Standard
> **Sc-1**	45	58926.0	1.1				mg/L	61029	Standard
Cl	35	32129.4	1.2				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	75514.4	2.1				ug/L	1547	Standard
P	31	20090.0	4.4				ug/L	15286	Standard
S	34	2098.5	7.9				ug/L	1572	Standard
Sr	88	1583.4	4.7				ug/L	73	Standard
C	12	130.0	7.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	561.0	7.3				mg/L	6	Standard
Ho-1	165	348.3	12.2				mg/L	10	Standard
Er	166	256.7	6.0				mg/L	13	Standard
I	127	84570.2	9.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		127.081	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.164	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098611
Report Date/Time: Wednesday, June 29, 2016 13:44:07
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.004
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.694
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098613
Sample Date/Time: Wednesday, June 29, 2016 13:45:02
Number of Replicates: 3
Autosampler Position: 323
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101811.1	2.5				ug/L	74699	Standard
	Be	9	780.0	6.8	**0.6573**	0.030	4.5	ug/L	8	Standard
	Al	27	69460342.3	2.9	**346.7492**	1.873	0.5	ug/L	1493	Standard
	Sc	45	66852.2	1.8				ug/L	61029	Standard
	Ti	47	16395.2	2.4	**40.4139**	0.750	1.9	ug/L	35	Standard
	V	51	44929.4	0.9	**5.3782**	0.102	1.9	ug/L	1435	Standard
	Cr	52	38038.0	1.4	**3.7847**	0.045	1.2	ug/L	7635	Standard
	Cr	53	5357.6	0.6	**5.1013**	0.082	1.6	ug/L	363	Standard
	Mn	55	1205027.8	2.1	**162.0424**	5.610	3.5	ug/L	747	Standard
	Co	59	15267.0	2.9	**2.1291**	0.080	3.8	ug/L	234	Standard
	Ni	60	13379.6	2.6	**6.5697**	0.065	1.0	ug/L	215	Standard
	Cu	65	74766.0	1.3	**37.5331**	0.708	1.9	ug/L	104	Standard
	Zn	66	182914.0	1.7	**156.4214**	2.327	1.5	ug/L	363	Standard
>	**Ge**	72	528583.9	2.0				ug/L	495142	Standard
	As	75	2539.4	2.2	**2.1832**	0.013	0.6	ug/L	-50	Standard
	Se	82	108.7	5.6	**0.7905**	0.071	9.0	ug/L	14	Standard
	Se-1	77	150.3	9.2	**1.3050**	0.161	12.3	ug/L	42	Standard
>	**Ga**	71	5017.5	0.9				mg/L	42	Standard
	Rb	85	95755.1	2.6				ug/L	17	Standard
	Y	89	484844.4	1.2				ug/L	417098	Standard
>	**Rh**	103	451.7	17.2				ug/L	13	Standard
	Mo	98	2711.0	1.7	**0.9675**	0.010	1.1	ug/L	11	Standard
	Ag	107	547.0	8.0	**0.0673**	0.005	7.7	ug/L	68	Standard
	Cd	111	100.9	10.4	**0.0383**	0.005	14.0	mg/L	2	Standard
	Cd	114	263.2	6.1	**0.0470**	0.003	5.8	ug/L	12	Standard
>	**In**	115	613966.8	2.2				ug/L	587441	Standard
	Sn	118	1991.8	3.6	**0.2317**	0.015	6.6	ug/L	497	Standard
	Sb	123	971.4	9.0	**0.1947**	0.014	7.1	ug/L	78	Standard
	Ba	135	97214.2	0.8	**45.4386**	0.665	1.5	ug/L	24	Standard
	Ce	140	163522.7	0.6				ug/L	25	Standard
>	**Tb**	159	1011575.1	2.5				ug/L	939701	Standard
	Ho	165	4682.4	4.2				ug/L	10	Standard
	Tl	203	573.7	3.0	**0.0642**	0.003	4.6	ug/L	4	Standard
	Tl	205	525.0	7.2	**0.0698**	0.005	7.8	ug/L	2	Standard
	Pb	206	15938.7	1.8	**2.7562**	0.018	0.6	ug/L	240	Standard
	Pb	207	13053.6	0.6	**2.5367**	0.043	1.7	ug/L	198	Standard
	Pb	208	53401.0	1.9	**2.6245**	0.013	0.5	ug/L	883	Standard
	U	238	38818.0	2.5	**5.3391**	0.057	1.1	ug/L	3	Standard
>	**Bi**	209	553002.2	2.3				ug/L	523984	Standard

Sample ID: L1606098613
Report Date/Time: Wednesday, June 29, 2016 13:47:19
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**88.0394**	31.556	35.8	mg/L	0	Standard
Mg	24	10193.4	4.3	**11.9448**	0.378	3.2	mg/L	37	Standard
K	39	798.4	2.0	**1.4974**	0.060	4.0	mg/L	38	Standard
Ca	43	1498.4	4.4	**83.9112**	4.508	5.4	mg/L	10	Standard
Fe	54	4251.3	3.0	**3.1535**	0.141	4.5	mg/L	179	Standard
Fe	57	1956.8	0.6	**4.9983**	0.081	1.6	mg/L	335	Standard
Sc-1	45	66852.2	1.8				mg/L	61029	Standard
Cl	35	38555.7	1.3				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	16405.6	4.1				ug/L	1547	Standard
P	31	22243.0	1.3				ug/L	15286	Standard
S	34	1961.8	3.3				ug/L	1572	Standard
Sr	88	2096.8	8.9				ug/L	73	Standard
C	12	170.0	21.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	107.9				mg/L	3	Standard
Dy	164	7687.7	5.1				mg/L	6	Standard
Ho-1	165	4682.4	4.2				mg/L	10	Standard
Er	166	4197.2	1.1				mg/L	13	Standard
I	127	34873.4	0.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		136.295	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		106.754	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098613
Report Date/Time: Wednesday, June 29, 2016 13:47:19
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.515
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.538
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098613
Report Date/Time: Wednesday, June 29, 2016 13:47:19
Page 3

Approved: July 04, 2016

Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098615
Sample Date/Time: Wednesday, June 29, 2016 13:48:13
Number of Replicates: 3
Autosampler Position: 324
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	104439.3	4.8				ug/L	74699	Standard
	Be	9	536.7	12.4	**0.4427**	0.037	8.4	ug/L	8	Standard
	Al	27	77862989.2	8.3	**378.5121**	14.291	3.8	ug/L	1493	Standard
	Sc	45	58982.2	11.3				ug/L	61029	Standard
	Ti	47	2461.2	11.7	**6.5990**	0.434	6.6	ug/L	35	Standard
	V	51	4945.2	8.0	**0.4817**	0.017	3.5	ug/L	1435	Standard
	Cr	52	13134.4	7.3	**0.8107**	0.065	8.0	ug/L	7635	Standard
	Cr	53	1810.1	8.3	**1.6977**	0.117	6.9	ug/L	363	Standard
	Mn	55	1145096.1	8.2	**169.9719**	4.564	2.7	ug/L	747	Standard
	Co	59	4019.2	11.3	**0.5913**	0.032	5.4	ug/L	234	Standard
	Ni	60	7177.1	13.9	**3.8310**	0.314	8.2	ug/L	215	Standard
	Cu	65	41403.8	8.7	**22.8999**	0.588	2.6	ug/L	104	Standard
	Zn	66	43831.2	7.7	**41.0562**	0.750	1.8	ug/L	363	Standard
>	**Ge**	72	478398.7	6.6				ug/L	495142	Standard
	As	75	638.2	9.5	**0.6495**	0.045	6.9	ug/L	-50	Standard
	Se	82	102.9	17.3	**0.8353**	0.172	20.6	ug/L	14	Standard
	Se-1	77	94.0	19.4	**0.7027**	0.185	26.4	ug/L	42	Standard
>	**Ga**	71	463.3	11.0				mg/L	42	Standard
	Rb	85	23055.9	7.7				ug/L	17	Standard
	Y	89	432649.5	7.6				ug/L	417098	Standard
>	**Rh**	103	1188.4	10.3				ug/L	13	Standard
	Mo	98	1350.9	6.9	**0.5044**	0.013	2.5	ug/L	11	Standard
	Ag	107	139.0	6.4	**0.0086**	0.002	19.3	ug/L	68	Standard
	Cd	111	44.6	6.8	**0.0137**	0.002	16.4	mg/L	2	Standard
	Cd	114	129.2	31.8	**0.0237**	0.010	43.4	ug/L	12	Standard
>	**In**	115	583863.5	7.1				ug/L	587441	Standard
	Sn	118	2513.5	5.1	**0.3387**	0.012	3.5	ug/L	497	Standard
	Sb	123	705.1	6.8	**0.1478**	0.002	1.6	ug/L	78	Standard
	Ba	135	26452.2	7.2	**12.9866**	0.138	1.1	ug/L	24	Standard
	Ce	140	27398.2	9.1				ug/L	25	Standard
>	**Tb**	159	958605.6	6.1				ug/L	939701	Standard
	Ho	165	645.0	4.0				ug/L	10	Standard
	Tl	203	134.0	9.5	**0.0146**	0.001	7.1	ug/L	4	Standard
	Tl	205	118.3	6.5	**0.0169**	0.002	13.2	ug/L	2	Standard
	Pb	206	2508.9	7.1	**0.4009**	0.002	0.4	ug/L	240	Standard
	Pb	207	1976.5	6.7	**0.3487**	0.007	2.0	ug/L	198	Standard
	Pb	208	8160.9	6.7	**0.3634**	0.004	1.2	ug/L	883	Standard
	U	238	17292.6	7.2	**2.5205**	0.006	0.2	ug/L	3	Standard
>	**Bi**	209	521380.1	7.3				ug/L	523984	Standard

Sample ID: L1606098615
Report Date/Time: Wednesday, June 29, 2016 13:50:30
Page 1

Approved: July 04, 2016

Na	23	15.0	57.7	**176.9361**	92.757	52.4	mg/L	0	Standard
Mg	24	28440.1	8.6	**38.0068**	1.323	3.5	mg/L	37	Standard
K	39	673.3	1.9	**1.4418**	0.213	14.8	mg/L	38	Standard
Ca	43	1596.8	5.0	**102.4296**	8.312	8.1	mg/L	10	Standard
Fe	54	1077.2	5.0	**0.8118**	0.070	8.7	mg/L	179	Standard
Fe	57	1203.4	10.1	**3.1714**	0.327	10.3	mg/L	335	Standard
> **Sc-1**	45	58982.2	11.3				mg/L	61029	Standard
Cl	35	41482.2	7.1				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	15648.1	7.3				ug/L	1547	Standard
P	31	20014.9	6.8				ug/L	15286	Standard
S	34	1913.5	9.5				ug/L	1572	Standard
Sr	88	5429.3	11.3				ug/L	73	Standard
C	12	203.3	27.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	1127.2	20.3				mg/L	6	Standard
Ho-1	165	645.0	4.0				mg/L	10	Standard
Er	166	546.7	6.4				mg/L	13	Standard
I	127	112969.0	6.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		139.814	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.618	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098615
Report Date/Time: Wednesday, June 29, 2016 13:50:30
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.391
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.503
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098615
Report Date/Time: Wednesday, June 29, 2016 13:50:30
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098617
Sample Date/Time: Wednesday, June 29, 2016 13:51:25
Number of Replicates: 3
Autosampler Position: 325
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95101.0	2.9				ug/L	74699	Standard
	Be	9	61.7	9.4	**0.0626**	0.004	5.8	ug/L	8	Standard
	Al	27	16761639.5	0.8	**89.6118**	1.975	2.2	ug/L	1493	Standard
	Sc	45	53828.8	0.6				ug/L	61029	Standard
	Ti	47	756.4	12.3	**2.0144**	0.253	12.6	ug/L	35	Standard
	V	51	3040.4	3.6	**0.2328**	0.013	5.5	ug/L	1435	Standard
	Cr	52	11892.0	0.9	**0.6842**	0.022	3.3	ug/L	7635	Standard
	Cr	53	1066.7	2.7	**0.8986**	0.036	4.1	ug/L	363	Standard
	Mn	55	57447.7	1.9	**8.6332**	0.089	1.0	ug/L	747	Standard
	Co	59	1276.7	2.2	**0.1685**	0.005	2.7	ug/L	234	Standard
	Ni	60	2114.5	5.4	**1.0778**	0.056	5.2	ug/L	215	Standard
	Cu	65	5578.0	1.5	**3.0859**	0.022	0.7	ug/L	104	Standard
	Zn	66	8715.8	0.2	**8.0380**	0.061	0.8	ug/L	363	Standard
>	**Ge**	72	465226.7	0.9				ug/L	495142	Standard
	As	75	386.7	5.7	**0.4269**	0.018	4.2	ug/L	-50	Standard
	Se	82	42.7	8.0	**0.2761**	0.036	13.2	ug/L	14	Standard
	Se-1	77	67.0	16.1	**0.3532**	0.150	42.5	ug/L	42	Standard
>	**Ga**	71	183.3	18.2				mg/L	42	Standard
	Rb	85	8882.6	2.7				ug/L	17	Standard
	Y	89	409246.0	2.8				ug/L	417098	Standard
>	**Rh**	103	295.0	13.6				ug/L	13	Standard
	Mo	98	1514.9	3.6	**0.5995**	0.027	4.5	ug/L	11	Standard
	Ag	107	86.3	3.5	**0.0012**	0.001	48.9	ug/L	68	Standard
	Cd	111	45.4	6.6	**0.0152**	0.002	11.0	mg/L	2	Standard
	Cd	114	119.3	31.2	**0.0227**	0.008	33.9	ug/L	12	Standard
>	**In**	115	551838.0	2.1				ug/L	587441	Standard
	Sn	118	896.7	9.3	**0.0688**	0.017	24.4	ug/L	497	Standard
	Sb	123	476.3	5.5	**0.1045**	0.008	7.8	ug/L	78	Standard
	Ba	135	51101.0	1.7	**26.5634**	0.162	0.6	ug/L	24	Standard
	Ce	140	4162.2	9.3				ug/L	25	Standard
>	**Tb**	159	892953.9	1.2				ug/L	939701	Standard
	Ho	165	111.7	20.2				ug/L	10	Standard
	Tl	203	71.0	9.8	**0.0072**	0.001	12.5	ug/L	4	Standard
	Tl	205	50.0	10.0	**0.0075**	0.001	10.0	ug/L	2	Standard
	Pb	206	834.7	3.9	**0.0918**	0.006	7.0	ug/L	240	Standard
	Pb	207	668.0	3.0	**0.0768**	0.004	5.0	ug/L	198	Standard
	Pb	208	2813.8	1.3	**0.0825**	0.002	2.7	ug/L	883	Standard
	U	238	2276.2	2.7	**0.3405**	0.008	2.4	ug/L	3	Standard
>	**Bi**	209	502274.9	0.4				ug/L	523984	Standard

Sample ID: L1606098617
Report Date/Time: Wednesday, June 29, 2016 13:53:42
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**21.6273**	37.451	173.2	mg/L	0	Standard
Mg	24	11037.4	4.7	**16.0863**	0.689	4.3	mg/L	37	Standard
K	39	470.0	10.1	**1.0642**	0.126	11.9	mg/L	38	Standard
Ca	43	383.3	8.5	**24.9281**	2.309	9.3	mg/L	10	Standard
Fe	54	307.3	17.1	**0.1571**	0.051	32.4	mg/L	179	Standard
Fe	57	490.0	5.4	**0.8143**	0.108	13.3	mg/L	335	Standard
> **Sc-1**	45	53828.8	0.6				mg/L	61029	Standard
Cl	35	29673.0	1.2				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	4564.0	9.2				ug/L	1547	Standard
P	31	17653.6	5.5				ug/L	15286	Standard
S	34	1876.8	5.1				ug/L	1572	Standard
Sr	88	1206.7	6.6				ug/L	73	Standard
C	12	186.7	18.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	173.0	22.9				mg/L	6	Standard
Ho-1	165	111.7	20.2				mg/L	10	Standard
Er	166	76.7	30.1				mg/L	13	Standard
I	127	8810.9	16.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		127.313	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.958	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098617
Report Date/Time: Wednesday, June 29, 2016 13:53:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		93.939
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.857
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098619
Sample Date/Time: Wednesday, June 29, 2016 13:54:36
Number of Replicates: 3
Autosampler Position: 326
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96013.3	1.6				ug/L	74699	Standard
	Be	9	225.0	13.5	**0.2062**	0.024	11.6	ug/L	8	Standard
	Al	27	3833929.3	1.4	**20.2847**	0.231	1.1	ug/L	1493	Standard
	Sc	45	56757.9	4.8				ug/L	61029	Standard
	Ti	47	657.0	3.0	**1.6551**	0.023	1.4	ug/L	35	Standard
	V	51	6468.0	2.1	**0.6763**	0.022	3.3	ug/L	1435	Standard
	Cr	52	12143.2	0.5	**0.6446**	0.043	6.7	ug/L	7635	Standard
	Cr	53	1533.4	8.3	**1.3595**	0.147	10.8	ug/L	363	Standard
	Mn	55	41319.9	1.1	**5.8936**	0.094	1.6	ug/L	747	Standard
	Co	59	1041.4	1.4	**0.1233**	0.005	3.8	ug/L	234	Standard
	Ni	60	1381.7	2.6	**0.6272**	0.001	0.2	ug/L	215	Standard
	Cu	65	22006.0	2.1	**11.9309**	0.100	0.8	ug/L	104	Standard
	Zn	66	12936.5	1.8	**11.6003**	0.101	0.9	ug/L	363	Standard
>	**Ge**	72	486496.6	2.4				ug/L	495142	Standard
	As	75	1429.3	4.7	**1.3578**	0.030	2.2	ug/L	-50	Standard
	Se	82	368.2	2.0	**3.2780**	0.104	3.2	ug/L	14	Standard
	Se-1	77	166.3	2.4	**1.6971**	0.106	6.2	ug/L	42	Standard
>	**Ga**	71	256.7	8.8				mg/L	42	Standard
	Rb	85	6678.1	5.5				ug/L	17	Standard
	Y	89	411615.1	2.5				ug/L	417098	Standard
>	**Rh**	103	45.0	40.1				ug/L	13	Standard
	Mo	98	2416.6	3.5	**0.9418**	0.055	5.8	ug/L	11	Standard
	Ag	107	106.0	14.6	**0.0040**	0.002	49.0	ug/L	68	Standard
	Cd	111	306.3	1.8	**0.1451**	0.002	1.4	mg/L	2	Standard
	Cd	114	658.6	9.3	**0.1319**	0.011	8.6	ug/L	12	Standard
>	**In**	115	562685.3	3.0				ug/L	587441	Standard
	Sn	118	1296.7	5.8	**0.1370**	0.009	6.5	ug/L	497	Standard
	Sb	123	1866.5	3.1	**0.4132**	0.007	1.8	ug/L	78	Standard
	Ba	135	49124.5	1.8	**25.0474**	0.324	1.3	ug/L	24	Standard
	Ce	140	4487.3	4.7				ug/L	25	Standard
>	**Tb**	159	930596.3	3.6				ug/L	939701	Standard
	Ho	165	203.3	14.2				ug/L	10	Standard
	Tl	203	145.3	15.0	**0.0163**	0.002	14.2	ug/L	4	Standard
	Tl	205	131.7	4.4	**0.0191**	0.000	1.8	ug/L	2	Standard
	Pb	206	1800.8	1.2	**0.2755**	0.007	2.5	ug/L	240	Standard
	Pb	207	1451.7	2.6	**0.2445**	0.002	0.8	ug/L	198	Standard
	Pb	208	5832.1	0.4	**0.2459**	0.009	3.7	ug/L	883	Standard
	U	238	34205.2	1.0	**5.1077**	0.154	3.0	ug/L	3	Standard
>	**Bi**	209	509656.0	3.1				ug/L	523984	Standard

Sample ID: L1606098619
Report Date/Time: Wednesday, June 29, 2016 13:56:53
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	92209.8	1.7	**128.1856**	8.381	6.5	mg/L	37	Standard
K	39	298.3	3.5	**0.5957**	0.033	5.6	mg/L	38	Standard
Ca	43	80.0	6.3	**2.9252**	0.551	18.8	mg/L	10	Standard
Fe	54	254.0	14.7	**0.0930**	0.030	32.4	mg/L	179	Standard
Fe	57	333.3	5.3	**0.1453**	0.073	50.0	mg/L	335	Standard
> **Sc-1**	45	56757.9	4.8				mg/L	61029	Standard
Cl	35	31152.7	1.5				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	53109.7	4.4				ug/L	1547	Standard
P	31	17403.3	2.0				ug/L	15286	Standard
S	34	1883.5	7.3				ug/L	1572	Standard
Sr	88	213.3	3.6				ug/L	73	Standard
C	12	256.7	12.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	288.4	16.8				mg/L	6	Standard
Ho-1	165	203.3	14.2				mg/L	10	Standard
Er	166	173.3	23.3				mg/L	13	Standard
I	127	470221.1	5.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		128.534	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.254	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098619
Report Date/Time: Wednesday, June 29, 2016 13:56:53
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	95.786
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	97.265
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098625
Sample Date/Time: Wednesday, June 29, 2016 13:57:48
Number of Replicates: 3
Autosampler Position: 327
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	102613.0	0.8				ug/L	74699	Standard
	Be	9	78.3	9.8	**0.0724**	0.006	8.8	ug/L	8	Standard
	Al	27	33905192.3	1.5	**167.9500**	3.413	2.0	ug/L	1493	Standard
	Sc	45	63494.4	0.7				ug/L	61029	Standard
	Ti	47	6149.2	2.1	**15.7581**	0.447	2.8	ug/L	35	Standard
	V	51	12969.6	1.6	**1.4837**	0.027	1.8	ug/L	1435	Standard
	Cr	52	17892.2	0.6	**1.3372**	0.031	2.3	ug/L	7635	Standard
	Cr	53	3548.7	0.7	**3.4248**	0.009	0.3	ug/L	363	Standard
	Mn	55	3105877.1	1.7	**436.1197**	10.576	2.4	ug/L	747	Standard
	Co	59	6022.9	2.6	**0.8547**	0.030	3.5	ug/L	234	Standard
	Ni	60	6318.7	1.1	**3.1764**	0.053	1.7	ug/L	215	Standard
	Cu	65	5119.5	1.0	**2.5860**	0.011	0.4	ug/L	104	Standard
	Zn	66	19836.6	0.9	**17.3044**	0.276	1.6	ug/L	363	Standard
>	**Ge**	72	506357.7	0.8				ug/L	495142	Standard
	As	75	1090.3	6.9	**1.0117**	0.071	7.0	ug/L	-50	Standard
	Se	82	334.6	3.6	**2.8439**	0.125	4.4	ug/L	14	Standard
	Se-1	77	270.0	7.0	**2.9935**	0.272	9.1	ug/L	42	Standard
>	**Ga**	71	1363.4	8.0				mg/L	42	Standard
	Rb	85	42725.2	0.9				ug/L	17	Standard
	Y	89	453938.8	1.4				ug/L	417098	Standard
>	**Rh**	103	198.3	19.1				ug/L	13	Standard
	Mo	98	1348.4	1.8	**0.4789**	0.005	1.1	ug/L	11	Standard
	Ag	107	259.3	4.5	**0.0252**	0.002	7.6	ug/L	68	Standard
	Cd	111	351.0	2.3	**0.1529**	0.002	1.6	mg/L	2	Standard
	Cd	114	808.7	5.9	**0.1490**	0.010	6.6	ug/L	12	Standard
>	**In**	115	613191.6	0.9				ug/L	587441	Standard
	Sn	118	795.0	1.9	**0.0357**	0.003	9.4	ug/L	497	Standard
	Sb	123	319.9	2.9	**0.0614**	0.002	3.0	ug/L	78	Standard
	Ba	135	112678.5	1.7	**52.7237**	0.702	1.3	ug/L	24	Standard
	Ce	140	37260.8	2.0				ug/L	25	Standard
>	**Tb**	159	1001354.7	0.4				ug/L	939701	Standard
	Ho	165	1051.7	3.8				ug/L	10	Standard
	Tl	203	322.7	6.7	**0.0353**	0.003	7.6	ug/L	4	Standard
	Tl	205	283.3	17.7	**0.0376**	0.006	16.6	ug/L	2	Standard
	Pb	206	8100.8	1.1	**1.3649**	0.025	1.8	ug/L	240	Standard
	Pb	207	6526.4	1.1	**1.2323**	0.027	2.2	ug/L	198	Standard
	Pb	208	27115.2	0.7	**1.2953**	0.018	1.4	ug/L	883	Standard
	U	238	26591.0	2.4	**3.6542**	0.116	3.2	ug/L	3	Standard
>	**Bi**	209	553363.8	1.0				ug/L	523984	Standard

Sample ID: L1606098625
Report Date/Time: Wednesday, June 29, 2016 14:00:05
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**55.1859**	55.233	100.1	mg/L	0	Standard
Mg	24	8027.1	0.4	**9.8947**	0.087	0.9	mg/L	37	Standard
K	39	523.3	14.3	**0.9970**	0.152	15.2	mg/L	38	Standard
Ca	43	861.7	11.2	**49.8050**	6.018	12.1	mg/L	10	Standard
Fe	54	1108.4	4.4	**0.7648**	0.035	4.5	mg/L	179	Standard
Fe	57	975.0	10.7	**2.1112**	0.341	16.2	mg/L	335	Standard
> **Sc-1**	45	63494.4	0.7				mg/L	61029	Standard
Cl	35	33805.7	1.3				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	35418.0	3.4				ug/L	1547	Standard
P	31	20113.3	1.1				ug/L	15286	Standard
S	34	2053.5	4.6				ug/L	1572	Standard
Sr	88	896.7	5.2				ug/L	73	Standard
C	12	130.0	35.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	1595.5	8.0				mg/L	6	Standard
Ho-1	165	1051.7	3.8				mg/L	10	Standard
Er	166	1000.0	4.0				mg/L	13	Standard
I	127	43091.5	8.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		137.369	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.265	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098625
Report Date/Time: Wednesday, June 29, 2016 14:00:05
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	104.384
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	105.607
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098625
Report Date/Time: Wednesday, June 29, 2016 14:00:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098627
Sample Date/Time: Wednesday, June 29, 2016 14:00:59
Number of Replicates: 3
Autosampler Position: 328
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98891.7	1.0				ug/L	74699	Standard
	Be	9	638.3	5.3	**0.5552**	0.024	4.3	ug/L	8	Standard
	Al	27	63228292.3	2.4	**324.9422**	4.583	1.4	ug/L	1493	Standard
	Sc	45	59185.4	3.1				ug/L	61029	Standard
	Ti	47	12099.8	1.5	**31.9570**	0.495	1.6	ug/L	35	Standard
	V	51	31610.6	0.8	**4.0089**	0.089	2.2	ug/L	1435	Standard
	Cr	52	29501.7	2.0	**2.9745**	0.066	2.2	ug/L	7635	Standard
	Cr	53	4110.6	4.5	**4.1359**	0.142	3.4	ug/L	363	Standard
	Mn	55	939161.3	0.8	**135.3408**	0.783	0.6	ug/L	747	Standard
	Co	59	11927.7	1.5	**1.7769**	0.013	0.7	ug/L	234	Standard
	Ni	60	10174.7	1.2	**5.3344**	0.025	0.5	ug/L	215	Standard
	Cu	65	63138.6	0.6	**33.9717**	0.415	1.2	ug/L	104	Standard
	Zn	66	96624.1	0.5	**88.3979**	1.021	1.2	ug/L	363	Standard
>	**Ge**	72	492984.7	1.4				ug/L	495142	Standard
	As	75	1814.9	1.7	**1.6870**	0.014	0.9	ug/L	-50	Standard
	Se	82	95.5	0.5	**0.7363**	0.016	2.2	ug/L	14	Standard
	Se-1	77	118.3	17.1	**1.0027**	0.256	25.5	ug/L	42	Standard
>	**Ga**	71	3340.4	1.3				mg/L	42	Standard
	Rb	85	72434.4	1.9				ug/L	17	Standard
	Y	89	474043.5	0.9				ug/L	417098	Standard
>	**Rh**	103	386.7	16.4				ug/L	13	Standard
	Mo	98	2396.5	0.7	**0.9091**	0.018	1.9	ug/L	11	Standard
	Ag	107	338.3	4.7	**0.0399**	0.003	6.7	ug/L	68	Standard
	Cd	111	101.2	9.9	**0.0414**	0.005	12.6	mg/L	2	Standard
	Cd	114	230.2	21.2	**0.0435**	0.009	20.9	ug/L	12	Standard
>	**In**	115	577417.6	1.2				ug/L	587441	Standard
	Sn	118	1793.4	6.6	**0.2178**	0.023	10.5	ug/L	497	Standard
	Sb	123	600.3	4.8	**0.1267**	0.008	6.1	ug/L	78	Standard
	Ba	135	70114.5	1.5	**34.8357**	0.414	1.2	ug/L	24	Standard
	Ce	140	125576.7	1.8				ug/L	25	Standard
>	**Tb**	159	953741.7	1.0				ug/L	939701	Standard
	Ho	165	3755.5	3.8				ug/L	10	Standard
	Tl	203	452.7	12.3	**0.0539**	0.006	11.6	ug/L	4	Standard
	Tl	205	385.0	10.1	**0.0548**	0.005	9.0	ug/L	2	Standard
	Pb	206	12183.9	0.5	**2.2443**	0.020	0.9	ug/L	240	Standard
	Pb	207	9920.2	0.8	**2.0519**	0.018	0.9	ug/L	198	Standard
	Pb	208	42107.7	1.6	**2.2055**	0.026	1.2	ug/L	883	Standard
	U	238	36138.4	1.7	**5.3249**	0.032	0.6	ug/L	3	Standard
>	**Bi**	209	516181.2	1.2				ug/L	523984	Standard

Sample ID: L1606098627
Report Date/Time: Wednesday, June 29, 2016 14:03:16
Page 1

Approved: July 04, 2016

Na	23	8.3	91.7	**99.8498**	90.493	90.6	mg/L	0	Standard
Mg	24	9524.7	1.9	**12.6179**	0.315	2.5	mg/L	37	Standard
K	39	723.4	10.8	**1.5367**	0.198	12.9	mg/L	38	Standard
Ca	43	1308.4	8.3	**82.6186**	4.814	5.8	mg/L	10	Standard
Fe	54	3102.8	2.2	**2.5749**	0.034	1.3	mg/L	179	Standard
Fe	57	1665.1	3.0	**4.7616**	0.022	0.5	mg/L	335	Standard
Sc-1	45	59185.4	3.1				mg/L	61029	Standard
Cl	35	38114.5	0.9				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	13185.7	5.6				ug/L	1547	Standard
P	31	20465.5	3.4				ug/L	15286	Standard
S	34	1960.1	7.7				ug/L	1572	Standard
Sr	88	1940.1	7.4				ug/L	73	Standard
C	12	216.7	23.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	3	Standard
Dy	164	5933.6	2.4				mg/L	6	Standard
Ho-1	165	3755.5	3.8				mg/L	10	Standard
Er	166	3430.4	0.6				mg/L	13	Standard
I	127	31850.1	1.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.387	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.564	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098627
Report Date/Time: Wednesday, June 29, 2016 14:03:16
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.294
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.511
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098627
Report Date/Time: Wednesday, June 29, 2016 14:03:16
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 14:04:12
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	95022.8	6.4				ug/L	74699	Standard
	Be	9	57413.6	2.5	**51.3575**	1.980	3.9	ug/L	8	Standard
	Al	27	7949888.2	1.3	**42.6051**	2.208	5.2	ug/L	1493	Standard
	Sc	45	58088.0	4.8				ug/L	61029	Standard
	Ti	47	37651.4	1.9	**100.2698**	3.249	3.2	ug/L	35	Standard
	V	51	404346.6	1.8	**53.8882**	0.844	1.6	ug/L	1435	Standard
	Cr	52	381698.1	1.1	**51.0223**	1.642	3.2	ug/L	7635	Standard
	Cr	53	47042.9	0.3	**51.1128**	1.842	3.6	ug/L	363	Standard
	Mn	55	357326.2	3.2	**51.6830**	0.417	0.8	ug/L	747	Standard
	Co	59	371699.6	2.5	**56.8155**	0.522	0.9	ug/L	234	Standard
	Ni	60	94541.7	1.4	**50.9011**	1.597	3.1	ug/L	215	Standard
	Cu	65	94983.8	1.5	**51.4773**	2.133	4.1	ug/L	104	Standard
	Zn	66	56900.0	0.7	**52.1829**	1.541	3.0	ug/L	363	Standard
>	Ge	72	490310.3	3.4				ug/L	495142	Standard
	As	75	57867.9	1.8	**52.2509**	0.921	1.8	ug/L	-50	Standard
	Se	82	5817.8	2.4	**53.4067**	0.847	1.6	ug/L	14	Standard
	Se-1	77	4012.5	1.4	**54.8814**	1.095	2.0	ug/L	42	Standard
>	Ga	71	80.0	6.3				mg/L	42	Standard
	Rb	85	831.7	8.5				ug/L	17	Standard
	Y	89	423000.5	2.4				ug/L	417098	Standard
>	Rh	103	26.7	57.3				ug/L	13	Standard
	Mo	98	279510.1	1.2	**103.0732**	1.481	1.4	ug/L	11	Standard
	Ag	107	337025.9	0.7	**50.8187**	0.653	1.3	ug/L	68	Standard
	Cd	111	104719.9	1.6	**49.2197**	0.492	1.0	mg/L	2	Standard
	Cd	114	254484.5	1.8	**48.7789**	1.357	2.8	ug/L	12	Standard
>	In	115	597695.8	1.6				ug/L	587441	Standard
	Sn	118	297589.5	3.3	**49.9660**	0.932	1.9	ug/L	497	Standard
	Sb	123	237800.4	0.6	**50.0611**	0.480	1.0	ug/L	78	Standard
	Ba	135	98785.4	1.3	**47.4247**	0.676	1.4	ug/L	24	Standard
	Ce	140	118.3	6.5				ug/L	25	Standard
>	Tb	159	962162.8	1.8				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	431721.7	1.4	**50.5906**	0.299	0.6	ug/L	4	Standard
	Tl	205	386667.7	1.8	**52.2055**	0.331	0.6	ug/L	2	Standard
	Pb	206	286495.7	0.6	**51.7458**	1.032	2.0	ug/L	240	Standard
	Pb	207	254256.8	1.2	**51.6830**	0.707	1.4	ug/L	198	Standard
	Pb	208	993731.8	1.4	**51.1193**	0.294	0.6	ug/L	883	Standard
	U	238	371822.3	1.0	**52.1805**	0.682	1.3	ug/L	3	Standard
>	Bi	209	542470.6	1.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 14:06:29
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**20.6892**	35.826	173.2	mg/L	0	Standard
Mg	24	4115.6	5.8	**5.5148**	0.066	1.2	mg/L	37	Standard
K	39	1971.8	1.5	**4.4711**	0.214	4.8	mg/L	38	Standard
Ca	43	71.7	16.1	**2.2502**	0.891	39.6	mg/L	10	Standard
Fe	54	5897.1	1.5	**5.1229**	0.272	5.3	mg/L	179	Standard
Fe	57	1798.4	2.6	**5.3592**	0.387	7.2	mg/L	335	Standard
Sc-1	45	58088.0	4.8				mg/L	61029	Standard
Cl	35	26443.8	4.7				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	1943.5	9.9				ug/L	1547	Standard
P	31	18503.0	1.3				ug/L	15286	Standard
S	34	2213.5	3.2				ug/L	1572	Standard
Sr	88	76.7	10.0				ug/L	73	Standard
C	12	120.0	43.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	2.9	217.0				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	3483.8	25.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	102.715		
Al	27	85.210		
Sc	45			
Ti	47	100.270		
V	51	107.776		
Cr	52	102.045		
Cr	53			
Mn	55	103.366		
Co	59	113.631		
Ni	60	101.802		
Cu	65	102.955		
Zn	66	104.366		
Ge	72		99.024	
As	75	104.502		
Se	82	106.813		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 14:06:29
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	103.073		
Ag	107	101.637		
Cd	111	98.439		
Cd	114			
In	115		101.746	
Sn	118	99.932		
Sb	123	100.122		
Ba	135	94.849		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.181		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.239		
U	238	104.361		
Bi	209		103.528	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 14:06:29
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 14:07:23
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87483.9	4.7				ug/L	74699	Standard
	Be	9	21.7	96.1	**0.0292**	0.021	73.0	ug/L	8	Standard
	Al	27	18230.8	94.8	**0.0950**	0.107	112.2	ug/L	1493	Standard
	Sc	45	56471.8	3.5				ug/L	61029	Standard
	Ti	47	36.0	46.7	**-0.0098**	0.050	510.8	ug/L	35	Standard
	V	51	1334.8	2.3	**-0.0125**	0.010	80.4	ug/L	1435	Standard
	Cr	52	6684.1	1.9	**-0.0897**	0.031	35.0	ug/L	7635	Standard
	Cr	53	393.3	6.4	**0.1115**	0.043	38.6	ug/L	363	Standard
	Mn	55	1346.4	57.0	**0.0546**	0.123	224.6	ug/L	747	Standard
	Co	59	278.3	17.3	**0.0065**	0.009	139.3	ug/L	234	Standard
	Ni	60	177.0	5.9	**-0.0261**	0.004	17.0	ug/L	215	Standard
	Cu	65	145.0	10.8	**-0.0230**	0.009	41.3	ug/L	104	Standard
	Zn	66	399.3	2.4	**-0.0757**	0.021	28.0	ug/L	363	Standard
>	**Ge**	72	478426.6	3.3				ug/L	495142	Standard
	As	75	-39.3	49.9	**0.0231**	0.019	81.6	ug/L	-50	Standard
	Se	82	16.5	11.7	**0.0181**	0.023	124.9	ug/L	14	Standard
	Se-1	77	38.7	26.9	**-0.0773**	0.131	170.1	ug/L	42	Standard
>	**Ga**	71	66.7	15.6				mg/L	42	Standard
	Rb	85	18.3	41.7				ug/L	17	Standard
	Y	89	416242.9	2.8				ug/L	417098	Standard
>	**Rh**	103	15.0	57.7				ug/L	13	Standard
	Mo	98	272.1	23.3	**0.0976**	0.024	24.4	ug/L	11	Standard
	Ag	107	106.0	6.8	**0.0036**	0.001	28.2	ug/L	68	Standard
	Cd	111	10.2	22.2	**-0.0030**	0.001	37.6	mg/L	2	Standard
	Cd	114	31.6	53.8	**0.0041**	0.003	80.7	ug/L	12	Standard
>	**In**	115	580179.3	0.7				ug/L	587441	Standard
	Sn	118	746.7	16.1	**0.0348**	0.022	62.5	ug/L	497	Standard
	Sb	123	368.4	24.4	**0.0758**	0.020	26.4	ug/L	78	Standard
	Ba	135	41.0	10.6	**0.0043**	0.002	51.4	ug/L	24	Standard
	Ce	140	38.3	61.6				ug/L	25	Standard
>	**Tb**	159	945404.9	0.5				ug/L	939701	Standard
	Ho	165	6.7	43.3				ug/L	10	Standard
	Tl	203	65.7	8.7	**0.0061**	0.001	8.8	ug/L	4	Standard
	Tl	205	95.0	45.9	**0.0133**	0.006	45.0	ug/L	2	Standard
	Pb	206	276.0	7.9	**-0.0202**	0.003	15.8	ug/L	240	Standard
	Pb	207	258.0	12.6	**-0.0163**	0.006	36.3	ug/L	198	Standard
	Pb	208	1134.3	12.6	**-0.0143**	0.008	56.7	ug/L	883	Standard
	U	238	91.0	75.8	**0.0087**	0.010	115.9	ug/L	3	Standard
>	**Bi**	209	530301.2	1.8				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 14:09:40
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	81.1	**0.0007**	0.054	7923.6	mg/L	37	Standard
K	39	40.0	45.1	**-0.0194**	0.039	202.6	mg/L	38	Standard
Ca	43	28.3	20.4	**-0.5841**	0.427	73.1	mg/L	10	Standard
Fe	54	129.2	39.5	**-0.0191**	0.049	258.2	mg/L	179	Standard
Fe	57	241.7	27.6	**-0.1803**	0.270	149.7	mg/L	335	Standard
> **Sc-1**	45	56471.8	3.5				mg/L	61029	Standard
Cl	35	26027.0	1.3				ug/L	20195	Standard
Kr	83	3.0	57.7				ug/L	4	Standard
Br	81	1783.4	6.5				ug/L	1547	Standard
P	31	18152.5	2.7				ug/L	15286	Standard
S	34	2168.5	1.2				ug/L	1572	Standard
Sr	88	56.7	27.0				ug/L	73	Standard
C	12	93.3	24.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	6.5	86.7				mg/L	6	Standard
Ho-1	165	6.7	43.3				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	2606.9	21.7				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		96.624	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 14:09:40
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	98.764	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	101.206	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 14:09:40
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098629
Sample Date/Time: Wednesday, June 29, 2016 14:10:36
Number of Replicates: 3
Autosampler Position: 329
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98303.0	2.6				ug/L	74699	Standard
	Be	9	480.0	19.8	**0.4216**	0.078	18.5	ug/L	8	Standard
	Al	27	74581592.0	2.0	**385.7852**	11.899	3.1	ug/L	1493	Standard
	Sc	45	57877.0	1.4				ug/L	61029	Standard
	Ti	47	6572.4	2.9	**18.4374**	0.321	1.7	ug/L	35	Standard
	V	51	8784.7	0.2	**1.0487**	0.019	1.8	ug/L	1435	Standard
	Cr	52	16164.0	1.7	**1.3092**	0.018	1.4	ug/L	7635	Standard
	Cr	53	2105.1	0.9	**2.1070**	0.052	2.5	ug/L	363	Standard
	Mn	55	1145494.2	1.1	**175.8506**	2.171	1.2	ug/L	747	Standard
	Co	59	6198.3	3.2	**0.9668**	0.035	3.6	ug/L	234	Standard
	Ni	60	8328.3	2.3	**4.6339**	0.069	1.5	ug/L	215	Standard
	Cu	65	39558.9	1.3	**22.6316**	0.139	0.6	ug/L	104	Standard
	Zn	66	37507.0	2.0	**36.2726**	0.401	1.1	ug/L	363	Standard
>	**Ge**	72	462898.1	1.3				ug/L	495142	Standard
	As	75	729.1	4.9	**0.7559**	0.035	4.6	ug/L	-50	Standard
	Se	82	97.9	9.3	**0.8168**	0.100	12.2	ug/L	14	Standard
	Se-1	77	92.7	14.7	**0.7365**	0.213	29.0	ug/L	42	Standard
>	**Ga**	71	1238.4	2.6				mg/L	42	Standard
	Rb	85	36981.8	1.6				ug/L	17	Standard
	Y	89	423750.9	0.9				ug/L	417098	Standard
>	**Rh**	103	1170.0	7.1				ug/L	13	Standard
	Mo	98	1333.1	1.3	**0.5111**	0.005	0.9	ug/L	11	Standard
	Ag	107	159.7	2.8	**0.0124**	0.001	8.3	ug/L	68	Standard
	Cd	111	50.7	12.0	**0.0171**	0.003	15.0	mg/L	2	Standard
	Cd	114	154.3	14.1	**0.0289**	0.004	13.0	ug/L	12	Standard
>	**In**	115	568573.6	2.1				ug/L	587441	Standard
	Sn	118	2385.2	8.8	**0.3269**	0.032	9.8	ug/L	497	Standard
	Sb	123	701.5	2.0	**0.1511**	0.004	2.9	ug/L	78	Standard
	Ba	135	49620.1	0.4	**25.0380**	0.447	1.8	ug/L	24	Standard
	Ce	140	69524.2	1.7				ug/L	25	Standard
>	**Tb**	159	941672.2	2.5				ug/L	939701	Standard
	Ho	165	1511.7	2.5				ug/L	10	Standard
	Tl	203	215.3	8.2	**0.0254**	0.002	6.7	ug/L	4	Standard
	Tl	205	171.7	4.4	**0.0251**	0.001	2.6	ug/L	2	Standard
	Pb	206	5244.6	2.1	**0.9485**	0.002	0.2	ug/L	240	Standard
	Pb	207	4168.9	1.9	**0.8423**	0.008	0.9	ug/L	198	Standard
	Pb	208	17565.0	0.7	**0.8990**	0.013	1.4	ug/L	883	Standard
	U	238	17410.3	1.3	**2.6228**	0.019	0.7	ug/L	3	Standard
>	**Bi**	209	504494.7	1.9				ug/L	523984	Standard

Sample ID: L1606098629
Report Date/Time: Wednesday, June 29, 2016 14:12:53
Page 1

Approved: July 04, 2016

Na	23	13.3	78.1	**161.6266**	126.012	78.0	mg/L	0	Standard
Mg	24	27959.1	0.6	**37.9939**	0.282	0.7	mg/L	37	Standard
K	39	733.4	5.5	**1.5946**	0.079	5.0	mg/L	38	Standard
Ca	43	1693.4	2.8	**110.2950**	4.030	3.7	mg/L	10	Standard
Fe	54	1507.3	3.7	**1.2090**	0.035	2.9	mg/L	179	Standard
Fe	57	1286.7	5.0	**3.5362**	0.177	5.0	mg/L	335	Standard
Sc-1	45	57877.0	1.4				mg/L	61029	Standard
Cl	35	39812.2	0.7				ug/L	20195	Standard
Kr	83	2.3	24.7				ug/L	4	Standard
Br	81	15855.0	0.8				ug/L	1547	Standard
P	31	19874.7	1.1				ug/L	15286	Standard
S	34	1861.8	8.3				ug/L	1572	Standard
Sr	88	5489.3	0.8				ug/L	73	Standard
C	12	143.3	29.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	2426.6	8.1				mg/L	6	Standard
Ho-1	165	1511.7	2.5				mg/L	10	Standard
Er	166	1466.7	15.1				mg/L	13	Standard
I	127	114010.2	6.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		131.599	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		93.488	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098629
Report Date/Time: Wednesday, June 29, 2016 14:12:53
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	96.788
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	96.280
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098629
Report Date/Time: Wednesday, June 29, 2016 14:12:53
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098639
Sample Date/Time: Wednesday, June 29, 2016 14:13:47
Number of Replicates: 3
Autosampler Position: 330
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95709.8	2.9				ug/L	74699	Standard
	Be	9	26.7	92.5	**0.0314**	0.022	70.9	ug/L	8	Standard
	Al	27	47777.1	19.3	**0.2396**	0.041	17.2	ug/L	1493	Standard
	Sc	45	57199.5	2.6				ug/L	61029	Standard
	Ti	47	140.3	60.7	**0.2633**	0.227	86.2	ug/L	35	Standard
	V	51	2113.6	5.5	**0.0856**	0.015	17.8	ug/L	1435	Standard
	Cr	52	11519.7	1.7	**0.5382**	0.024	4.4	ug/L	7635	Standard
	Cr	53	731.7	8.2	**0.4658**	0.068	14.5	ug/L	363	Standard
	Mn	55	10046.3	0.4	**1.3019**	0.003	0.2	ug/L	747	Standard
	Co	59	277.3	4.9	**0.0049**	0.002	38.1	ug/L	234	Standard
	Ni	60	403.7	2.5	**0.0928**	0.006	6.6	ug/L	215	Standard
	Cu	65	2594.2	1.0	**1.2969**	0.011	0.8	ug/L	104	Standard
	Zn	66	5488.0	1.5	**4.5940**	0.065	1.4	ug/L	363	Standard
>	**Ge**	72	492746.6	0.5				ug/L	495142	Standard
	As	75	-23.9	159.7	**0.0381**	0.034	90.0	ug/L	-50	Standard
	Se	82	16.9	42.3	**0.0162**	0.065	398.3	ug/L	14	Standard
	Se-1	77	44.7	15.2	**-0.0080**	0.091	1136.4	ug/L	42	Standard
>	**Ga**	71	35.0	24.7				mg/L	42	Standard
	Rb	85	228.3	6.7				ug/L	17	Standard
	Y	89	426313.2	2.4				ug/L	417098	Standard
>	**Rh**	103	20.0	66.1				ug/L	13	Standard
	Mo	98	146.2	8.6	**0.0487**	0.005	10.8	ug/L	11	Standard
	Ag	107	84.7	10.6	**-0.0000**	0.001	3427.7	ug/L	68	Standard
	Cd	111	14.4	26.2	**-0.0011**	0.002	163.5	mg/L	2	Standard
	Cd	114	36.0	7.5	**0.0048**	0.001	11.2	ug/L	12	Standard
>	**In**	115	592771.3	1.1				ug/L	587441	Standard
	Sn	118	1206.7	2.0	**0.1101**	0.007	6.0	ug/L	497	Standard
	Sb	123	2747.6	2.7	**0.5790**	0.016	2.8	ug/L	78	Standard
	Ba	135	327.7	1.8	**0.1426**	0.002	1.3	ug/L	24	Standard
	Ce	140	215.0	46.7				ug/L	25	Standard
>	**Tb**	159	942434.8	1.8				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	30.7	28.5	**0.0019**	0.001	54.2	ug/L	4	Standard
	Tl	205	31.7	59.8	**0.0045**	0.003	57.0	ug/L	2	Standard
	Pb	206	347.7	8.5	**-0.0077**	0.005	63.4	ug/L	240	Standard
	Pb	207	296.0	3.8	**-0.0091**	0.002	22.4	ug/L	198	Standard
	Pb	208	1207.0	5.0	**-0.0113**	0.003	23.3	ug/L	883	Standard
	U	238	16.7	21.1	**-0.0021**	0.000	23.4	ug/L	3	Standard
>	**Bi**	209	536687.4	0.8				ug/L	523984	Standard

Sample ID: L1606098639
Report Date/Time: Wednesday, June 29, 2016 14:16:04
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	78.3	9.8	**0.0427**	0.008	19.8	mg/L	37	Standard
K	39	33.3	56.8	**-0.0347**	0.047	136.1	mg/L	38	Standard
Ca	43	38.3	27.2	**0.0551**	0.672	1218.0	mg/L	10	Standard
Fe	54	152.5	13.6	**-0.0001**	0.021	24429.9	mg/L	179	Standard
Fe	57	270.0	10.3	**-0.0961**	0.083	86.0	mg/L	335	Standard
Sc-1	45	57199.5	2.6				mg/L	61029	Standard
Cl	35	30633.6	0.6				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	2093.5	12.2				ug/L	1547	Standard
P	31	17446.7	2.4				ug/L	15286	Standard
S	34	1971.8	2.9				ug/L	1572	Standard
Sr	88	48.3	26.0				ug/L	73	Standard
C	12	116.7	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	9.0	108.0				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	4304.0	50.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		128.128	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		99.516	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098639
Report Date/Time: Wednesday, June 29, 2016 14:16:04
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.907
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.424
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098641
Sample Date/Time: Wednesday, June 29, 2016 14:16:59
Number of Replicates: 3
Autosampler Position: 331
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87267.4	3.2				ug/L	74699	Standard
	Be	9	11.7	49.5	**0.0189**	0.005	28.8	ug/L	8	Standard
	Al	27	605511.0	1.7	**3.5149**	0.083	2.4	ug/L	1493	Standard
	Sc	45	54390.8	2.9				ug/L	61029	Standard
	Ti	47	93.0	12.4	**0.1442**	0.032	22.4	ug/L	35	Standard
	V	51	1794.1	5.9	**0.0496**	0.015	30.2	ug/L	1435	Standard
	Cr	52	10365.2	1.5	**0.4195**	0.024	5.7	ug/L	7635	Standard
	Cr	53	613.3	8.2	**0.3551**	0.058	16.5	ug/L	363	Standard
	Mn	55	4513.0	1.6	**0.5208**	0.012	2.3	ug/L	747	Standard
	Co	59	310.3	5.3	**0.0112**	0.002	21.0	ug/L	234	Standard
	Ni	60	490.0	6.7	**0.1461**	0.017	11.6	ug/L	215	Standard
	Cu	65	1018.0	2.6	**0.4608**	0.017	3.7	ug/L	104	Standard
	Zn	66	3968.5	2.5	**3.2944**	0.091	2.8	ug/L	363	Standard
>	**Ge**	72	479837.7	0.4				ug/L	495142	Standard
	As	75	-18.8	39.3	**0.0423**	0.007	15.9	ug/L	-50	Standard
	Se	82	16.1	20.6	**0.0139**	0.032	229.1	ug/L	14	Standard
	Se-1	77	51.0	14.8	**0.0980**	0.104	105.7	ug/L	42	Standard
>	**Ga**	71	46.7	22.3				mg/L	42	Standard
	Rb	85	2146.8	6.0				ug/L	17	Standard
	Y	89	405082.7	1.5				ug/L	417098	Standard
>	**Rh**	103	10.0	86.6				ug/L	13	Standard
	Mo	98	183.9	8.1	**0.0672**	0.006	9.5	ug/L	11	Standard
	Ag	107	85.0	8.5	**0.0009**	0.001	123.2	ug/L	68	Standard
	Cd	111	58.3	9.8	**0.0215**	0.003	12.3	mg/L	2	Standard
	Cd	114	162.0	10.0	**0.0312**	0.003	9.6	ug/L	12	Standard
>	**In**	115	556306.9	1.6				ug/L	587441	Standard
	Sn	118	983.4	6.5	**0.0831**	0.013	16.0	ug/L	497	Standard
	Sb	123	290.8	14.6	**0.0617**	0.011	17.2	ug/L	78	Standard
	Ba	135	1301.4	2.0	**0.6554**	0.007	1.1	ug/L	24	Standard
	Ce	140	1251.7	2.7				ug/L	25	Standard
>	**Tb**	159	894371.7	2.0				ug/L	939701	Standard
	Ho	165	23.3	61.9				ug/L	10	Standard
	Tl	203	26.7	21.3	**0.0015**	0.001	46.3	ug/L	4	Standard
	Tl	205	18.3	68.6	**0.0028**	0.002	63.1	ug/L	2	Standard
	Pb	206	368.7	5.0	**-0.0008**	0.004	496.5	ug/L	240	Standard
	Pb	207	301.3	5.5	**-0.0052**	0.004	82.3	ug/L	198	Standard
	Pb	208	1198.4	1.5	**-0.0089**	0.002	20.6	ug/L	883	Standard
	U	238	7.7	32.8	**-0.0033**	0.000	11.3	ug/L	3	Standard
>	**Bi**	209	513924.4	1.5				ug/L	523984	Standard

Sample ID: L1606098641
Report Date/Time: Wednesday, June 29, 2016 14:19:15
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	638.3	13.6	**0.8624**	0.151	17.5	mg/L	37	Standard
K	39	86.7	3.3	**0.1010**	0.012	12.3	mg/L	38	Standard
Ca	43	31.7	48.2	**-0.2871**	1.074	374.0	mg/L	10	Standard
Fe	54	129.1	21.5	**-0.0150**	0.029	192.4	mg/L	179	Standard
Fe	57	250.0	12.2	**-0.1181**	0.141	119.7	mg/L	335	Standard
Sc-1	45	54390.8	2.9				mg/L	61029	Standard
Cl	35	28123.4	3.1				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1706.8	4.2				ug/L	1547	Standard
P	31	17456.7	1.2				ug/L	15286	Standard
S	34	1960.1	4.3				ug/L	1572	Standard
Sr	88	86.7	3.3				ug/L	73	Standard
C	12	86.7	35.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	22.9	67.6				mg/L	6	Standard
Ho-1	165	23.3	61.9				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	6631.5	5.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		116.826	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.909	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098641
Report Date/Time: Wednesday, June 29, 2016 14:19:15
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.700
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.080
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102602
Sample Date/Time: Wednesday, June 29, 2016 14:51:20
Number of Replicates: 3
Autosampler Position: 332
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98959.1	4.1				ug/L	74699	Standard
	Be	9	893.4	7.1	**0.7730**	0.024	3.1	ug/L	8	Standard
	Al	27	86838575.1	3.0	**446.1622**	5.675	1.3	ug/L	1493	Standard
	Sc	45	59841.3	1.2				ug/L	61029	Standard
	Ti	47	724.7	2.7	**1.9675**	0.086	4.4	ug/L	35	Standard
	V	51	-99401.9	27.0	**-14.4927**	3.863	26.7	ug/L	1435	Standard
	Cr	52	14286817.6	0.7	**2092.1702**	38.595	1.8	ug/L	7635	Standard
	Cr	53	1819264.9	1.0	**2136.8631**	30.527	1.4	ug/L	363	Standard
	Mn	55	9369393.8	2.3	**1460.7209**	18.166	1.2	ug/L	747	Standard
	Co	59	953653.0	0.8	**156.7618**	2.325	1.5	ug/L	234	Standard
	Ni	60	10094931.6	0.5	**5854.0954**	89.919	1.5	ug/L	215	Standard
	Cu	65	555284.9	0.4	**323.8508**	6.370	2.0	ug/L	104	Standard
	Zn	66	64094.6	1.0	**63.2545**	1.058	1.7	ug/L	363	Standard
>	**Ge**	72	456115.5	1.9				ug/L	495142	Standard
	As	75	3627.2	1.9	**3.5767**	0.135	3.8	ug/L	-50	Standard
	Se	82	1754.2	3.0	**17.2111**	0.289	1.7	ug/L	14	Standard
	Se-1	77	3450.1	4.6	**50.7001**	3.400	6.7	ug/L	42	Standard
>	**Ga**	71	493.3	9.8				mg/L	42	Standard
	Rb	85	15579.7	1.3				ug/L	17	Standard
	Y	89	510259.3	1.5				ug/L	417098	Standard
>	**Rh**	103	536.7	10.8				ug/L	13	Standard
	Mo	98	52406.3	2.2	**22.0721**	0.200	0.9	ug/L	11	Standard
	Ag	107	797.7	2.2	**0.1246**	0.004	3.1	ug/L	68	Standard
	Cd	111	4820.0	1.8	**2.5808**	0.025	1.0	mg/L	2	Standard
	Cd	114	11654.3	2.8	**2.5494**	0.039	1.5	ug/L	12	Standard
>	**In**	115	523117.8	1.3				ug/L	587441	Standard
	Sn	118	2368.5	5.4	**0.3606**	0.023	6.4	ug/L	497	Standard
	Sb	123	2240.5	1.3	**0.5347**	0.001	0.2	ug/L	78	Standard
	Ba	135	135783.1	0.9	**74.4848**	0.361	0.5	ug/L	24	Standard
	Ce	140	104687.1	1.1				ug/L	25	Standard
>	**Tb**	159	898543.7	0.7				ug/L	939701	Standard
	Ho	165	6684.8	2.2				ug/L	10	Standard
	Tl	203	178.0	4.1	**0.0225**	0.001	5.6	ug/L	4	Standard
	Tl	205	203.3	37.2	**0.0321**	0.012	38.4	ug/L	2	Standard
	Pb	206	1304.4	5.0	**0.2026**	0.016	8.1	ug/L	240	Standard
	Pb	207	1055.7	2.5	**0.1793**	0.004	2.5	ug/L	198	Standard
	Pb	208	4341.2	2.0	**0.1854**	0.004	2.0	ug/L	883	Standard
	U	238	4585.0	0.8	**0.7423**	0.007	0.9	ug/L	3	Standard
>	**Bi**	209	467454.5	1.3				ug/L	523984	Standard

Sample ID: L1606102602
Report Date/Time: Wednesday, June 29, 2016 14:53:37
Page 1

Approved: July 04, 2016

Na	23	5.0	0.0	**58.7812**	0.729	1.2	mg/L	0	Standard
Mg	24	188890.8	2.9	**248.6881**	10.353	4.2	mg/L	37	Standard
K	39	2185.2	4.4	**4.8106**	0.175	3.6	mg/L	38	Standard
Ca	43	1135.0	6.5	**70.5642**	3.873	5.5	mg/L	10	Standard
Fe	54	9878.2	4.8	**8.4031**	0.399	4.8	mg/L	179	Standard
Fe	57	3267.0	5.5	**10.2558**	0.748	7.3	mg/L	335	Standard
> **Sc-1**	45	59841.3	1.2				mg/L	61029	Standard
Cl	35	28240.3	1.1				ug/L	20195	Standard
Kr	83	2.7	142.0				ug/L	4	Standard
Br	81	339155.5	2.6				ug/L	1547	Standard
P	31	29607.2	0.7				ug/L	15286	Standard
S	34	2378.5	1.9				ug/L	1572	Standard
Sr	88	1895.1	7.4				ug/L	73	Standard
C	12	120.0	28.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	86.7	35.3				mg/L	3	Standard
Dy	164	9299.8	2.7				mg/L	6	Standard
Ho-1	165	6684.8	2.2				mg/L	10	Standard
Er	166	6331.3	1.7				mg/L	13	Standard
I	127	153619.9	12.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.478	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.118	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606102602
Report Date/Time: Wednesday, June 29, 2016 14:53:37
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	89.050
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	89.212
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
V 51 Lower	V	51	

Sample ID: L1606102602
Report Date/Time: Wednesday, June 29, 2016 14:53:37
Page 3

Approved: July 04, 2016

Cr 52 Upper, S, EEE	Cr	52
Cr 53 Upper, S, EEE	Cr	53
Mn 55 Upper, S, EEE	Mn	55
Co 59 Upper, S, EEE	Co	59
Ni 60 Upper, S, EEE	Ni	60
Cu 65 Upper, S, EEE	Cu	65

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102604
Sample Date/Time: Wednesday, June 29, 2016 14:54:31
Number of Replicates: 3
Autosampler Position: 333
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82386.8	1.6				ug/L	74699	Standard
	Be	9	41.7	50.0	**0.0505**	0.021	41.9	ug/L	8	Standard
	Al	27	6418180.7	1.0	**39.5946**	1.017	2.6	ug/L	1493	Standard
	Sc	45	68191.6	3.4				ug/L	61029	Standard
	Ti	47	2431.5	5.8	**6.2763**	0.263	4.2	ug/L	35	Standard
	V	51	8932.1	2.8	**0.9822**	0.034	3.4	ug/L	1435	Standard
	Cr	52	19971.2	5.8	**1.6585**	0.113	6.8	ug/L	7635	Standard
	Cr	53	16018.5	1.2	**16.9336**	0.479	2.8	ug/L	363	Standard
	Mn	55	134583.1	0.5	**19.1035**	0.326	1.7	ug/L	747	Standard
	Co	59	5374.0	2.6	**0.7730**	0.008	1.1	ug/L	234	Standard
	Ni	60	7430.2	8.6	**3.8249**	0.274	7.2	ug/L	215	Standard
	Cu	65	1488.1	5.9	**0.6921**	0.035	5.0	ug/L	104	Standard
	Zn	66	5451.0	0.8	**4.5156**	0.102	2.3	ug/L	363	Standard
>	**Ge**	72	497254.0	1.7				ug/L	495142	Standard
	As	75	427.8	16.4	**0.4399**	0.063	14.3	ug/L	-50	Standard
	Se	82	161.6	5.9	**1.3278**	0.065	4.9	ug/L	14	Standard
	Se-1	77	695.3	4.1	**8.8552**	0.225	2.5	ug/L	42	Standard
>	**Ga**	71	420.0	2.4				mg/L	42	Standard
	Rb	85	8292.3	2.6				ug/L	17	Standard
	Y	89	430205.9	2.4				ug/L	417098	Standard
>	**Rh**	103	40.0	12.5				ug/L	13	Standard
	Mo	98	177.9	8.2	**0.0624**	0.005	8.3	ug/L	11	Standard
	Ag	107	97.3	5.7	**0.0023**	0.001	38.0	ug/L	68	Standard
	Cd	111	330.0	5.3	**0.1533**	0.008	5.4	mg/L	2	Standard
	Cd	114	777.4	4.0	**0.1527**	0.007	4.3	ug/L	12	Standard
>	**In**	115	575297.3	2.0				ug/L	587441	Standard
	Sn	118	1836.8	3.5	**0.2263**	0.005	2.3	ug/L	497	Standard
	Sb	123	541.3	2.5	**0.1143**	0.005	4.2	ug/L	78	Standard
	Ba	135	259403.2	1.0	**129.4148**	1.250	1.0	ug/L	24	Standard
	Ce	140	48153.1	1.6				ug/L	25	Standard
>	**Tb**	159	953820.3	0.6				ug/L	939701	Standard
	Ho	165	763.4	3.9				ug/L	10	Standard
	Tl	203	147.7	11.5	**0.0162**	0.002	12.5	ug/L	4	Standard
	Tl	205	135.0	25.9	**0.0192**	0.005	26.2	ug/L	2	Standard
	Pb	206	2333.8	1.6	**0.3679**	0.004	1.0	ug/L	240	Standard
	Pb	207	1897.5	1.5	**0.3318**	0.006	1.8	ug/L	198	Standard
	Pb	208	7911.8	1.2	**0.3500**	0.006	1.7	ug/L	883	Standard
	U	238	272.0	3.5	**0.0353**	0.002	4.3	ug/L	3	Standard
>	**Bi**	209	521315.2	0.9				ug/L	523984	Standard

Sample ID: L1606102604
Report Date/Time: Wednesday, June 29, 2016 14:56:48
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	45454.8	3.6	**52.4459**	0.539	1.0	mg/L	37	Standard
K	39	88.3	3.3	**0.0609**	0.011	18.4	mg/L	38	Standard
Ca	43	100.0	8.7	**3.1440**	0.659	21.0	mg/L	10	Standard
Fe	54	556.2	7.6	**0.2841**	0.037	12.9	mg/L	179	Standard
Fe	57	410.0	16.9	**0.1734**	0.212	122.1	mg/L	335	Standard
> **Sc-1**	45	68191.6	3.4				mg/L	61029	Standard
Cl	35	26725.6	1.1				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	35244.3	3.6				ug/L	1547	Standard
P	31	35015.4	1.7				ug/L	15286	Standard
S	34	2091.8	5.9				ug/L	1572	Standard
Sr	88	225.0	20.0				ug/L	73	Standard
C	12	103.3	39.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	93.3	12.4				mg/L	3	Standard
Dy	164	1034.1	12.7				mg/L	6	Standard
Ho-1	165	763.4	3.9				mg/L	10	Standard
Er	166	680.0	15.4				mg/L	13	Standard
I	127	80476.5	1.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.292	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.426	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606102604
Report Date/Time: Wednesday, June 29, 2016 14:56:48
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.933
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.491
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Ba 135 Upper, S, EEE	Ba	135	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102606
Sample Date/Time: Wednesday, June 29, 2016 14:57:42
Number of Replicates: 3
Autosampler Position: 334
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93356.7	0.9				ug/L	74699	Standard
	Be	9	136.7	43.6	**0.1318**	0.054	40.9	ug/L	8	Standard
	Al	27	4727947.0	2.8	**25.7324**	0.870	3.4	ug/L	1493	Standard
	Sc	45	71265.6	1.4				ug/L	61029	Standard
	Ti	47	7167.7	1.1	**18.5395**	0.116	0.6	ug/L	35	Standard
	V	51	33631.6	0.4	**4.1963**	0.093	2.2	ug/L	1435	Standard
	Cr	52	43684.3	3.9	**4.7878**	0.170	3.6	ug/L	7635	Standard
	Cr	53	18915.1	0.4	**19.8540**	0.330	1.7	ug/L	363	Standard
	Mn	55	527300.2	1.1	**74.5358**	0.480	0.6	ug/L	747	Standard
	Co	59	47125.2	1.5	**6.9996**	0.016	0.2	ug/L	234	Standard
	Ni	60	57578.9	1.0	**30.2025**	0.274	0.9	ug/L	215	Standard
	Cu	65	3787.5	1.8	**1.9028**	0.007	0.4	ug/L	104	Standard
	Zn	66	29686.7	1.5	**26.3448**	0.262	1.0	ug/L	363	Standard
>	**Ge**	72	502153.4	1.7				ug/L	495142	Standard
	As	75	1552.5	5.8	**1.4257**	0.058	4.0	ug/L	-50	Standard
	Se	82	328.6	4.9	**2.8136**	0.101	3.6	ug/L	14	Standard
	Se-1	77	866.0	3.6	**11.0672**	0.235	2.1	ug/L	42	Standard
>	**Ga**	71	2006.8	3.5				mg/L	42	Standard
	Rb	85	28431.7	2.8				ug/L	17	Standard
	Y	89	467809.8	1.5				ug/L	417098	Standard
>	**Rh**	103	28.3	36.7				ug/L	13	Standard
	Mo	98	2788.7	2.4	**1.0189**	0.018	1.8	ug/L	11	Standard
	Ag	107	187.0	8.5	**0.0152**	0.002	14.7	ug/L	68	Standard
	Cd	111	184.5	8.6	**0.0785**	0.007	8.5	mg/L	2	Standard
	Cd	114	502.9	12.6	**0.0940**	0.013	13.9	ug/L	12	Standard
>	**In**	115	599816.0	1.2				ug/L	587441	Standard
	Sn	118	1385.1	2.5	**0.1375**	0.003	2.2	ug/L	497	Standard
	Sb	123	315.5	7.5	**0.0620**	0.005	7.5	ug/L	78	Standard
	Ba	135	37805.8	2.1	**18.0728**	0.162	0.9	ug/L	24	Standard
	Ce	140	83749.2	1.1				ug/L	25	Standard
>	**Tb**	159	971524.4	1.2				ug/L	939701	Standard
	Ho	165	1393.4	2.1				ug/L	10	Standard
	Tl	203	289.7	7.1	**0.0339**	0.002	6.5	ug/L	4	Standard
	Tl	205	258.3	11.0	**0.0368**	0.004	11.7	ug/L	2	Standard
	Pb	206	3853.2	1.2	**0.6604**	0.004	0.6	ug/L	240	Standard
	Pb	207	3010.0	1.5	**0.5732**	0.009	1.6	ug/L	198	Standard
	Pb	208	13337.3	2.3	**0.6473**	0.013	1.9	ug/L	883	Standard
	U	238	859.0	4.4	**0.1221**	0.005	4.0	ug/L	3	Standard
>	**Bi**	209	516635.2	0.9				ug/L	523984	Standard

Sample ID: L1606102606
Report Date/Time: Wednesday, June 29, 2016 14:59:59
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.3563**	28.321	173.2	mg/L	0	Standard
Mg	24	103702.1	2.0	**114.5988**	3.552	3.1	mg/L	37	Standard
K	39	251.7	12.0	**0.3627**	0.062	17.2	mg/L	38	Standard
Ca	43	61.7	26.1	**0.8208**	0.922	112.4	mg/L	10	Standard
Fe	54	1412.4	2.3	**0.8871**	0.015	1.7	mg/L	179	Standard
Fe	57	645.0	2.8	**0.8030**	0.026	3.3	mg/L	335	Standard
> **Sc-1**	45	71265.6	1.4				mg/L	61029	Standard
Cl	35	28407.3	0.4				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	49253.3	3.6				ug/L	1547	Standard
P	31	41409.9	2.5				ug/L	15286	Standard
S	34	2068.5	4.4				ug/L	1572	Standard
Sr	88	155.0	22.6				ug/L	73	Standard
C	12	133.3	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	57.3				mg/L	3	Standard
Dy	164	2353.9	2.8				mg/L	6	Standard
Ho-1	165	1393.4	2.1				mg/L	10	Standard
Er	166	1453.4	9.2				mg/L	13	Standard
I	127	3067238.7	11.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		124.978	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.416	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606102606
Report Date/Time: Wednesday, June 29, 2016 14:59:59
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	102.107	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	98.597	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102607
Sample Date/Time: Wednesday, June 29, 2016 15:00:54
Number of Replicates: 3
Autosampler Position: 335
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86025.2	1.8				ug/L	74699	Standard
	Be	9	83.3	29.6	**0.0901**	0.026	28.3	ug/L	8	Standard
	Al	27	3430436.9	1.2	**20.2582**	0.272	1.3	ug/L	1493	Standard
	Sc	45	67818.2	2.7				ug/L	61029	Standard
	Ti	47	2650.2	2.1	**6.1960**	0.095	1.5	ug/L	35	Standard
	V	51	9706.1	1.4	**0.9634**	0.025	2.6	ug/L	1435	Standard
	Cr	52	14237.0	2.0	**0.7086**	0.035	4.9	ug/L	7635	Standard
	Cr	53	2521.9	11.8	**2.1295**	0.275	12.9	ug/L	363	Standard
	Mn	55	50752.2	3.1	**6.4246**	0.167	2.6	ug/L	747	Standard
	Co	59	8541.7	4.1	**1.1290**	0.040	3.6	ug/L	234	Standard
	Ni	60	2905.6	5.8	**1.2753**	0.074	5.8	ug/L	215	Standard
	Cu	65	1384.7	3.1	**0.5672**	0.017	3.0	ug/L	104	Standard
	Zn	66	5145.2	1.6	**3.7928**	0.039	1.0	ug/L	363	Standard
>	**Ge**	72	549096.7	0.7				ug/L	495142	Standard
	As	75	307.3	24.2	**0.3069**	0.059	19.1	ug/L	-50	Standard
	Se	82	131.2	8.7	**0.9395**	0.088	9.3	ug/L	14	Standard
	Se-1	77	177.7	11.5	**1.5705**	0.241	15.4	ug/L	42	Standard
>	**Ga**	71	765.0	18.4				mg/L	42	Standard
	Rb	85	11574.4	1.2				ug/L	17	Standard
	Y	89	476523.2	1.2				ug/L	417098	Standard
>	**Rh**	103	40.0	54.5				ug/L	13	Standard
	Mo	98	552.0	3.2	**0.1876**	0.008	4.1	ug/L	11	Standard
	Ag	107	103.3	5.6	**0.0019**	0.001	51.8	ug/L	68	Standard
	Cd	111	131.7	9.5	**0.0509**	0.006	12.0	mg/L	2	Standard
	Cd	114	303.7	9.9	**0.0531**	0.005	10.1	ug/L	12	Standard
>	**In**	115	629575.3	1.6				ug/L	587441	Standard
	Sn	118	1851.8	1.5	**0.2011**	0.002	0.9	ug/L	497	Standard
	Sb	123	475.2	2.5	**0.0908**	0.004	4.2	ug/L	78	Standard
	Ba	135	119108.2	2.3	**54.2912**	1.481	2.7	ug/L	24	Standard
	Ce	140	65755.7	1.8				ug/L	25	Standard
>	**Tb**	159	1004571.7	2.5				ug/L	939701	Standard
	Ho	165	1471.7	8.7				ug/L	10	Standard
	Tl	203	123.3	5.2	**0.0121**	0.001	4.7	ug/L	4	Standard
	Tl	205	98.3	10.6	**0.0129**	0.002	11.8	ug/L	2	Standard
	Pb	206	2854.9	1.5	**0.4229**	0.014	3.4	ug/L	240	Standard
	Pb	207	2325.8	2.5	**0.3831**	0.018	4.6	ug/L	198	Standard
	Pb	208	9906.6	0.9	**0.4144**	0.011	2.7	ug/L	883	Standard
	U	238	660.3	1.2	**0.0843**	0.002	2.3	ug/L	3	Standard
>	**Bi**	209	566874.3	1.4				ug/L	523984	Standard

Sample ID: L1606102607
Report Date/Time: Wednesday, June 29, 2016 15:03:11
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	9059.4	5.1	**10.4668**	0.674	6.4	mg/L	37	Standard
K	39	175.0	7.6	**0.2337**	0.017	7.4	mg/L	38	Standard
Ca	43	76.7	10.0	**1.8315**	0.359	19.6	mg/L	10	Standard
Fe	54	778.8	6.8	**0.4564**	0.046	10.1	mg/L	179	Standard
Fe	57	450.0	7.8	**0.3039**	0.129	42.4	mg/L	335	Standard
> **Sc-1**	45	67818.2	2.7				mg/L	61029	Standard
Cl	35	27265.2	1.0				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	4797.4	2.1				ug/L	1547	Standard
P	31	22328.1	2.9				ug/L	15286	Standard
S	34	1868.4	11.0				ug/L	1572	Standard
Sr	88	170.0	23.0				ug/L	73	Standard
C	12	116.7	4.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	2276.8	3.2				mg/L	6	Standard
Ho-1	165	1471.7	8.7				mg/L	10	Standard
Er	166	1323.4	4.9				mg/L	13	Standard
I	127	142556.3	31.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		115.163	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.897	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606102607
Report Date/Time: Wednesday, June 29, 2016 15:03:11
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		107.172
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.185
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 15:04:07
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	83434.4	4.1				ug/L	74699	Standard
	Be	9	52151.3	1.6	**53.0799**	1.438	2.7	ug/L	8	Standard
	Al	27	8001829.5	1.2	**48.7762**	1.563	3.2	ug/L	1493	Standard
	Sc	45	63472.7	2.1				ug/L	61029	Standard
	Ti	47	41544.5	0.5	**100.8598**	1.507	1.5	ug/L	35	Standard
	V	51	430823.0	0.6	**52.3478**	0.567	1.1	ug/L	1435	Standard
	Cr	52	414183.8	1.0	**50.4520**	0.020	0.0	ug/L	7635	Standard
	Cr	53	51435.5	2.1	**50.9262**	0.597	1.2	ug/L	363	Standard
	Mn	55	402509.0	2.2	**53.1108**	1.647	3.1	ug/L	747	Standard
	Co	59	395807.1	0.7	**55.1753**	0.902	1.6	ug/L	234	Standard
	Ni	60	102515.7	1.1	**50.3079**	0.215	0.4	ug/L	215	Standard
	Cu	65	102297.5	1.1	**50.5232**	0.685	1.4	ug/L	104	Standard
	Zn	66	61288.7	1.4	**51.2272**	0.860	1.7	ug/L	363	Standard
>	Ge	72	537577.7	1.0				ug/L	495142	Standard
	As	75	61234.0	0.5	**50.4168**	0.738	1.5	ug/L	-50	Standard
	Se	82	6132.0	1.1	**51.3300**	1.089	2.1	ug/L	14	Standard
	Se-1	77	4299.6	0.9	**53.6055**	0.977	1.8	ug/L	42	Standard
>	Ga	71	78.3	42.5				mg/L	42	Standard
	Rb	85	866.7	7.4				ug/L	17	Standard
	Y	89	460058.1	1.7				ug/L	417098	Standard
>	Rh	103	33.3	31.2				ug/L	13	Standard
	Mo	98	295506.5	0.5	**103.9241**	2.114	2.0	ug/L	11	Standard
	Ag	107	348188.0	0.5	**50.0657**	0.748	1.5	ug/L	68	Standard
	Cd	111	111062.0	1.3	**49.7777**	0.346	0.7	mg/L	2	Standard
	Cd	114	271614.5	0.9	**49.6432**	0.990	2.0	ug/L	12	Standard
>	In	115	626816.0	1.9				ug/L	587441	Standard
	Sn	118	313548.1	1.3	**50.2246**	1.384	2.8	ug/L	497	Standard
	Sb	123	246409.0	0.1	**49.4692**	0.867	1.8	ug/L	78	Standard
	Ba	135	105775.6	0.8	**48.4328**	1.255	2.6	ug/L	24	Standard
	Ce	140	66.7	48.8				ug/L	25	Standard
>	Tb	159	996778.4	1.8				ug/L	939701	Standard
	Ho	165	16.7	34.6				ug/L	10	Standard
	Tl	203	442598.6	0.7	**50.9461**	0.764	1.5	ug/L	4	Standard
	Tl	205	402531.2	1.7	**53.3816**	0.705	1.3	ug/L	2	Standard
	Pb	206	287928.0	0.2	**51.0713**	0.393	0.8	ug/L	240	Standard
	Pb	207	257921.1	0.2	**51.4941**	0.566	1.1	ug/L	198	Standard
	Pb	208	1040092.9	0.4	**52.5571**	0.545	1.0	ug/L	883	Standard
	U	238	381264.5	1.2	**52.5495**	0.479	0.9	ug/L	3	Standard
>	Bi	209	552262.9	0.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:06:24
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4305.6	5.5	**5.2771**	0.213	4.0	mg/L	37	Standard
K	39	2275.2	2.4	**4.7239**	0.217	4.6	mg/L	38	Standard
Ca	43	126.7	42.6	**5.1579**	3.228	62.6	mg/L	10	Standard
Fe	54	6415.2	4.5	**5.0956**	0.328	6.4	mg/L	179	Standard
Fe	57	1851.8	4.1	**4.9827**	0.373	7.5	mg/L	335	Standard
> **Sc-1**	45	63472.7	2.1				mg/L	61029	Standard
Cl	35	28119.4	1.3				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2310.2	13.5				ug/L	1547	Standard
P	31	19991.5	1.1				ug/L	15286	Standard
S	34	2273.5	5.1				ug/L	1572	Standard
Sr	88	71.7	10.7				ug/L	73	Standard
C	12	66.7	8.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	8.6	116.8				mg/L	6	Standard
Ho-1	165	16.7	34.6				mg/L	10	Standard
Er	166	30.0	0.0				mg/L	13	Standard
I	127	25116.6	23.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	106.160		
Al	27	97.552		
Sc	45			
Ti	47	100.860		
V	51	104.696		
Cr	52	100.904		
Cr	53			
Mn	55	106.222		
Co	59	110.351		
Ni	60	100.616		
Cu	65	101.046		
Zn	66	102.454		
> Ge	72		108.570	
As	75	100.834		
Se	82	102.660		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:06:24
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	103.924		
Ag	107	100.131		
Cd	111	99.555		
Cd	114			
In	115			106.703
Sn	118	100.449		
Sb	123	98.938		
Ba	135	96.866		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.892		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.114		
U	238	105.099		
Bi	209			105.397
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:06:24
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 15:07:18
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	85012.9	2.0				ug/L	74699	Standard
	Be	9	28.3	27.0	**0.0359**	0.008	21.1	ug/L	8	Standard
	Al	27	3885.6	39.7	**0.0100**	0.010	97.9	ug/L	1493	Standard
	Sc	45	66216.1	1.0				ug/L	61029	Standard
	Ti	47	42.3	11.2	**-0.0066**	0.011	167.9	ug/L	35	Standard
	V	51	1610.8	4.5	**0.0001**	0.010	18188.0	ug/L	1435	Standard
	Cr	52	7733.0	4.2	**-0.0667**	0.042	62.7	ug/L	7635	Standard
	Cr	53	656.7	7.6	**0.3218**	0.049	15.4	ug/L	363	Standard
	Mn	55	934.4	2.8	**-0.0250**	0.004	15.9	ug/L	747	Standard
	Co	59	270.3	16.3	**0.0003**	0.006	2278.1	ug/L	234	Standard
	Ni	60	353.0	46.0	**0.0492**	0.080	162.9	ug/L	215	Standard
	Cu	65	160.7	5.6	**-0.0245**	0.004	18.2	ug/L	104	Standard
	Zn	66	427.3	1.4	**-0.0957**	0.006	6.2	ug/L	363	Standard
>	Ge	72	539852.0	0.5				ug/L	495142	Standard
	As	75	-35.2	72.3	**0.0308**	0.021	67.1	ug/L	-50	Standard
	Se	82	23.4	13.4	**0.0577**	0.027	47.1	ug/L	14	Standard
	Se-1	77	65.0	15.6	**0.1944**	0.131	67.4	ug/L	42	Standard
>	Ga	71	48.3	15.8				mg/L	42	Standard
	Rb	85	13.3	21.7				ug/L	17	Standard
	Y	89	460334.1	1.4				ug/L	417098	Standard
>	Rh	103	15.0	57.7				ug/L	13	Standard
	Mo	98	325.6	30.7	**0.1063**	0.035	33.0	ug/L	11	Standard
	Ag	107	121.0	14.9	**0.0041**	0.003	62.9	ug/L	68	Standard
	Cd	111	13.2	31.2	**-0.0021**	0.002	85.0	mg/L	2	Standard
	Cd	114	51.9	43.2	**0.0072**	0.004	56.5	ug/L	12	Standard
>	In	115	641477.2	0.6				ug/L	587441	Standard
	Sn	118	860.0	23.5	**0.0400**	0.031	77.0	ug/L	497	Standard
	Sb	123	396.3	20.9	**0.0735**	0.016	21.4	ug/L	78	Standard
	Ba	135	45.3	20.4	**0.0043**	0.004	99.1	ug/L	24	Standard
	Ce	140	36.7	15.7				ug/L	25	Standard
>	Tb	159	1015779.3	0.6				ug/L	939701	Standard
	Ho	165	5.0	100.0				ug/L	10	Standard
	Tl	203	76.0	35.5	**0.0066**	0.003	43.8	ug/L	4	Standard
	Tl	205	63.3	18.2	**0.0082**	0.001	17.5	ug/L	2	Standard
	Pb	206	309.7	4.7	**-0.0185**	0.002	12.2	ug/L	240	Standard
	Pb	207	263.0	6.6	**-0.0197**	0.003	14.7	ug/L	198	Standard
	Pb	208	1081.0	8.8	**-0.0217**	0.004	19.1	ug/L	883	Standard
	U	238	55.3	39.0	**0.0029**	0.003	97.3	ug/L	3	Standard
>	Bi	209	576716.4	0.8				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:09:35
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.8612**	30.928	173.2	mg/L	0	Standard
Mg	24	48.3	39.2	**-0.0075**	0.023	309.1	mg/L	37	Standard
K	39	65.0	38.5	**0.0181**	0.050	275.9	mg/L	38	Standard
Ca	43	31.7	9.1	**-0.6795**	0.172	25.4	mg/L	10	Standard
Fe	54	165.7	15.0	**-0.0089**	0.019	210.6	mg/L	179	Standard
Fe	57	258.3	12.9	**-0.2654**	0.099	37.5	mg/L	335	Standard
Sc-1	45	66216.1	1.0				mg/L	61029	Standard
Cl	35	27856.9	0.3				ug/L	20195	Standard
Kr	83	4.0	43.3				ug/L	4	Standard
Br	81	1983.5	3.1				ug/L	1547	Standard
P	31	20308.6	3.1				ug/L	15286	Standard
S	34	2271.8	3.7				ug/L	1572	Standard
Sr	88	51.7	22.3				ug/L	73	Standard
C	12	73.3	34.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	20.0	86.6				mg/L	6	Standard
Ho-1	165	5.0	100.0				mg/L	10	Standard
Er	166	0.0					mg/L	13	Standard
I	127	13137.4	11.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		109.030	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:09:35
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.199
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.064
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:09:35
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBS M5 WG573099-02
Sample Date/Time: Wednesday, June 29, 2016 15:10:32
Number of Replicates: 3
Autosampler Position: 336
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82837.5	1.4				ug/L	74699	Standard
	Be	9	20.0	75.0	**0.0280**	0.015	53.9	ug/L	8	Standard
	Al	27	15988.5	12.5	**0.0848**	0.014	16.0	ug/L	1493	Standard
	Sc	45	64671.1	2.4				ug/L	61029	Standard
	Ti	47	45.0	20.0	**-0.0015**	0.023	1562.0	ug/L	35	Standard
	V	51	1647.0	1.8	**0.0017**	0.004	236.7	ug/L	1435	Standard
	Cr	52	11218.5	4.3	**0.3456**	0.052	15.0	ug/L	7635	Standard
	Cr	53	1063.4	5.0	**0.7116**	0.068	9.6	ug/L	363	Standard
	Mn	55	1756.1	5.1	**0.0801**	0.011	14.2	ug/L	747	Standard
	Co	59	312.3	6.3	**0.0055**	0.003	48.1	ug/L	234	Standard
	Ni	60	1377.4	15.3	**0.5415**	0.099	18.4	ug/L	215	Standard
	Cu	65	261.7	8.1	**0.0235**	0.011	45.2	ug/L	104	Standard
	Zn	66	1613.1	3.7	**0.8814**	0.040	4.6	ug/L	363	Standard
>	**Ge**	72	547319.5	1.5				ug/L	495142	Standard
	As	75	-42.5	100.6	**0.0255**	0.034	133.0	ug/L	-50	Standard
	Se	82	18.0	18.7	**0.0100**	0.027	272.2	ug/L	14	Standard
	Se-1	77	65.0	16.6	**0.1841**	0.145	78.6	ug/L	42	Standard
>	**Ga**	71	66.7	22.9				mg/L	42	Standard
	Rb	85	36.7	56.8				ug/L	17	Standard
	Y	89	450480.7	0.9				ug/L	417098	Standard
>	**Rh**	103	26.7	10.8				ug/L	13	Standard
	Mo	98	139.4	1.7	**0.0435**	0.001	2.2	ug/L	11	Standard
	Ag	107	102.7	18.1	**0.0019**	0.003	140.0	ug/L	68	Standard
	Cd	111	13.1	71.9	**-0.0020**	0.004	209.6	mg/L	2	Standard
	Cd	114	24.5	136.0	**0.0024**	0.006	256.9	ug/L	12	Standard
>	**In**	115	624818.4	0.5				ug/L	587441	Standard
	Sn	118	765.0	2.8	**0.0284**	0.003	12.0	ug/L	497	Standard
	Sb	123	189.1	20.4	**0.0339**	0.008	23.4	ug/L	78	Standard
	Ba	135	83.0	7.9	**0.0221**	0.003	14.0	ug/L	24	Standard
	Ce	140	301.7	18.8				ug/L	25	Standard
>	**Tb**	159	992078.8	0.6				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	40.0	57.5	**0.0027**	0.003	94.2	ug/L	4	Standard
	Tl	205	43.3	29.0	**0.0058**	0.002	28.5	ug/L	2	Standard
	Pb	206	319.7	11.1	**-0.0160**	0.006	36.6	ug/L	240	Standard
	Pb	207	262.0	4.6	**-0.0191**	0.002	12.7	ug/L	198	Standard
	Pb	208	1094.7	2.8	**-0.0202**	0.002	8.1	ug/L	883	Standard
	U	238	22.0	58.0	**-0.0015**	0.002	114.9	ug/L	3	Standard
>	**Bi**	209	568214.1	0.6				ug/L	523984	Standard

Sample ID: PBS M5 WG573099-02
Report Date/Time: Wednesday, June 29, 2016 15:12:48
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	43.3	17.6	**-0.0124**	0.008	67.7	mg/L	37	Standard
K	39	35.0	51.5	**-0.0416**	0.036	86.5	mg/L	38	Standard
Ca	43	30.0	28.9	**-0.7286**	0.543	74.5	mg/L	10	Standard
Fe	54	174.1	18.8	**0.0014**	0.029	2030.6	mg/L	179	Standard
Fe	57	266.7	22.0	**-0.2161**	0.204	94.4	mg/L	335	Standard
> **Sc-1**	45	64671.1	2.4				mg/L	61029	Standard
Cl	35	28617.0	1.0				ug/L	20195	Standard
Kr	83	2.0	100.0				ug/L	4	Standard
Br	81	1953.5	6.4				ug/L	1547	Standard
P	31	19132.1	3.4				ug/L	15286	Standard
S	34	2278.5	6.0				ug/L	1572	Standard
Sr	88	68.3	29.6				ug/L	73	Standard
C	12	83.3	13.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.5	39.0				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	23.3	49.5				mg/L	13	Standard
I	127	8569.1	14.4				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		110.895	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		110.538	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBS M5 WG573099-02
Report Date/Time: Wednesday, June 29, 2016 15:12:48
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.363
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.441
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBS M5 WG573099-02
Report Date/Time: Wednesday, June 29, 2016 15:12:48
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSS M5 WG573099-03
Sample Date/Time: Wednesday, June 29, 2016 15:13:43
Number of Replicates: 3
Autosampler Position: 337
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83551.5	2.3				ug/L	74699	Standard
	Be	9	25751.9	2.9	**26.1593**	0.514	2.0	ug/L	8	Standard
	Al	27	17496.8	6.9	**0.0930**	0.006	6.2	ug/L	1493	Standard
	Sc	45	66115.6	1.8				ug/L	61029	Standard
	Ti	47	52.7	14.3	**0.0175**	0.016	92.6	ug/L	35	Standard
	V	51	220707.2	1.1	**26.4464**	0.390	1.5	ug/L	1435	Standard
	Cr	52	221818.0	0.7	**26.2660**	0.546	2.1	ug/L	7635	Standard
	Cr	53	27359.7	0.9	**26.6591**	0.406	1.5	ug/L	363	Standard
	Mn	55	204818.9	1.5	**26.6717**	0.451	1.7	ug/L	747	Standard
	Co	59	196776.0	1.4	**27.1279**	0.304	1.1	ug/L	234	Standard
	Ni	60	55651.4	0.4	**26.9765**	0.492	1.8	ug/L	215	Standard
	Cu	65	54836.4	0.8	**26.7594**	0.556	2.1	ug/L	104	Standard
	Zn	66	32934.1	1.1	**27.0387**	0.809	3.0	ug/L	363	Standard
>	**Ge**	72	543210.6	2.2				ug/L	495142	Standard
	As	75	31720.5	1.5	**25.8765**	0.351	1.4	ug/L	-50	Standard
	Se	82	3103.9	1.9	**25.6428**	0.383	1.5	ug/L	14	Standard
	Se-1	77	2231.2	2.7	**27.2211**	0.278	1.0	ug/L	42	Standard
>	**Ga**	71	55.0	47.2				mg/L	42	Standard
	Rb	85	35.0	42.9				ug/L	17	Standard
	Y	89	467716.3	0.8				ug/L	417098	Standard
>	**Rh**	103	13.3	43.3				ug/L	13	Standard
	Mo	98	194.8	10.9	**0.0603**	0.008	12.8	ug/L	11	Standard
	Ag	107	10349.2	1.8	**1.4201**	0.038	2.7	ug/L	68	Standard
	Cd	111	56717.0	1.5	**24.4699**	0.504	2.1	mg/L	2	Standard
	Cd	114	136634.8	1.7	**24.0382**	0.506	2.1	ug/L	12	Standard
>	**In**	115	651063.9	0.8				ug/L	587441	Standard
	Sn	118	761.7	5.5	**0.0229**	0.007	28.6	ug/L	497	Standard
	Sb	123	128047.0	1.0	**24.7441**	0.436	1.8	ug/L	78	Standard
	Ba	135	54350.3	0.9	**23.9455**	0.395	1.6	ug/L	24	Standard
	Ce	140	61.7	26.1				ug/L	25	Standard
>	**Tb**	159	1033159.0	0.5				ug/L	939701	Standard
	Ho	165	16.7	17.3				ug/L	10	Standard
	Tl	203	227266.1	0.5	**24.7175**	0.218	0.9	ug/L	4	Standard
	Tl	205	203485.9	1.6	**25.5015**	0.511	2.0	ug/L	2	Standard
	Pb	206	152382.5	0.6	**25.5050**	0.255	1.0	ug/L	240	Standard
	Pb	207	130131.8	0.7	**24.5130**	0.243	1.0	ug/L	198	Standard
	Pb	208	530327.1	0.9	**25.2835**	0.262	1.0	ug/L	883	Standard
	U	238	189200.1	1.2	**24.6397**	0.334	1.4	ug/L	3	Standard
>	**Bi**	209	584433.1	0.4				ug/L	523984	Standard

Sample ID: LCSS M5 WG573099-03
Report Date/Time: Wednesday, June 29, 2016 15:16:00
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	41.7	36.7	**-0.0156**	0.017	112.1	mg/L	37	Standard
K	39	38.3	52.7	**-0.0354**	0.042	119.1	mg/L	38	Standard
Ca	43	25.0	52.9	**-1.0705**	0.758	70.8	mg/L	10	Standard
Fe	54	141.4	21.3	**-0.0274**	0.025	92.5	mg/L	179	Standard
Fe	57	241.7	15.7	**-0.3151**	0.126	40.1	mg/L	335	Standard
Sc-1	45	66115.6	1.8				mg/L	61029	Standard
Cl	35	28385.2	0.6				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	2056.8	1.5				ug/L	1547	Standard
P	31	18911.8	2.1				ug/L	15286	Standard
S	34	2256.8	4.9				ug/L	1572	Standard
Sr	88	70.0	39.8				ug/L	73	Standard
C	12	60.0	66.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	9.5	5.0				mg/L	6	Standard
Ho-1	165	16.7	17.3				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	6019.5	14.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		111.851	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		109.708	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: LCSS M5 WG573099-03
Report Date/Time: Wednesday, June 29, 2016 15:16:00
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.831
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.536
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSS M5 WG573099-03
Report Date/Time: Wednesday, June 29, 2016 15:16:00
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088803 WG573099-01
Sample Date/Time: Wednesday, June 29, 2016 15:16:55
Number of Replicates: 3
Autosampler Position: 338
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87864.2	2.8				ug/L	74699	Standard
	Be	9	1871.8	7.3	**1.8142**	0.100	5.5	ug/L	8	Standard
	Al	27	17464438.5	2.2	**101.0382**	1.686	1.7	ug/L	1493	Standard
	Sc	45	88916.8	1.0				ug/L	61029	Standard
	Ti	47	8105.2	1.6	**20.2503**	0.364	1.8	ug/L	35	Standard
	V	51	266665.9	0.8	**33.4185**	0.188	0.6	ug/L	1435	Standard
	Cr	52	308197.4	0.8	**38.5684**	0.415	1.1	ug/L	7635	Standard
	Cr	53	37658.4	2.2	**38.4602**	0.742	1.9	ug/L	363	Standard
	Mn	55	19789542.3	0.8	**2705.9402**	31.717	1.2	ug/L	747	Standard
	Co	59	276278.5	0.6	**39.7944**	0.401	1.0	ug/L	234	Standard
	Ni	60	160372.4	1.4	**81.4241**	1.275	1.6	ug/L	215	Standard
	Cu	65	107769.0	1.4	**55.0229**	0.902	1.6	ug/L	104	Standard
	Zn	66	213527.8	0.7	**185.6475**	1.406	0.8	ug/L	363	Standard
>	**Ge**	72	520092.2	0.4				ug/L	495142	Standard
	As	75	6170.4	1.0	**5.3042**	0.070	1.3	ug/L	-50	Standard
	Se	82	90.8	2.7	**0.6494**	0.023	3.5	ug/L	14	Standard
	Se-1	77	157.7	4.9	**1.4330**	0.106	7.4	ug/L	42	Standard
>	**Ga**	71	47343.9	1.1				mg/L	42	Standard
	Rb	85	380728.2	1.5				ug/L	17	Standard
	Y	89	1106618.3	0.2				ug/L	417098	Standard
>	**Rh**	103	63.3	24.1				ug/L	13	Standard
	Mo	98	980.4	3.2	**0.3366**	0.010	3.0	ug/L	11	Standard
	Ag	107	955.0	6.8	**0.1236**	0.011	9.1	ug/L	68	Standard
	Cd	111	2225.7	2.0	**0.9829**	0.017	1.8	mg/L	2	Standard
	Cd	114	5479.9	1.8	**0.9924**	0.001	0.1	ug/L	12	Standard
>	**In**	115	631161.9	1.7				ug/L	587441	Standard
	Sn	118	413.3	14.8	**-0.0289**	0.009	31.0	ug/L	497	Standard
	Sb	123	370.7	34.4	**0.0695**	0.025	35.4	ug/L	78	Standard
	Ba	135	385428.4	1.0	**175.2729**	2.367	1.4	ug/L	24	Standard
	Ce	140	2007459.4	3.6				ug/L	25	Standard
>	**Tb**	159	1048778.2	1.5				ug/L	939701	Standard
	Ho	165	53168.1	1.9				ug/L	10	Standard
	Tl	203	2332.8	1.5	**0.2668**	0.007	2.6	ug/L	4	Standard
	Tl	205	2056.8	1.1	**0.2730**	0.001	0.5	ug/L	2	Standard
	Pb	206	102474.4	0.3	**18.1326**	0.189	1.0	ug/L	240	Standard
	Pb	207	82544.6	1.5	**16.4326**	0.141	0.9	ug/L	198	Standard
	Pb	208	353310.3	0.4	**17.8060**	0.180	1.0	ug/L	883	Standard
	U	238	9355.9	0.7	**1.2854**	0.017	1.3	ug/L	3	Standard
>	**Bi**	209	552227.6	1.4				ug/L	523984	Standard

Sample ID: L1606088803 WG573099-01
Report Date/Time: Wednesday, June 29, 2016 15:19:11
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**13.3403**	23.097	173.1	mg/L	0	Standard
Mg	24	320.0	5.6	**0.2185**	0.016	7.5	mg/L	37	Standard
K	39	1080.0	10.5	**1.5233**	0.156	10.2	mg/L	38	Standard
Ca	43	916.7	12.0	**37.2112**	4.625	12.4	mg/L	10	Standard
Fe	54	75887.5	2.1	**44.0256**	1.050	2.4	mg/L	179	Standard
Fe	57	19502.5	1.2	**44.4256**	0.652	1.5	mg/L	335	Standard
> **Sc-1**	45	88916.8	1.0				mg/L	61029	Standard
Cl	35	27513.6	0.5				ug/L	20195	Standard
Kr	83	0.3	173.2				ug/L	4	Standard
Br	81	3467.1	2.9				ug/L	1547	Standard
P	31	21865.8	2.4				ug/L	15286	Standard
S	34	2631.9	4.0				ug/L	1572	Standard
Sr	88	220.0	6.0				ug/L	73	Standard
C	12	66.7	37.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	36.7	15.7				mg/L	3	Standard
Dy	164	86599.4	2.5				mg/L	6	Standard
Ho-1	165	53168.1	1.9				mg/L	10	Standard
Er	166	47260.3	0.7				mg/L	13	Standard
I	127	6809.9	9.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		117.625	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		105.039	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606088803 WG573099-01
Report Date/Time: Wednesday, June 29, 2016 15:19:11
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		107.443
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.390
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606088803 WG573099-01
Report Date/Time: Wednesday, June 29, 2016 15:19:11
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088803S WG573099-04
Sample Date/Time: Wednesday, June 29, 2016 15:20:06
Number of Replicates: 3
Autosampler Position: 339
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	88784.5	2.9				ug/L	74699	Standard
	Be	9	27508.3	0.4	**26.3080**	0.631	2.4	ug/L	8	Standard
	Al	27	20794150.5	3.6	**119.0173**	0.902	0.8	ug/L	1493	Standard
	Sc	45	96992.9	2.3				ug/L	61029	Standard
	Ti	47	11606.4	2.4	**28.2553**	0.997	3.5	ug/L	35	Standard
	V	51	594819.7	3.2	**72.7593**	3.186	4.4	ug/L	1435	Standard
	Cr	52	602098.1	1.8	**74.2282**	2.325	3.1	ug/L	7635	Standard
	Cr	53	74295.1	1.5	**74.1206**	1.783	2.4	ug/L	363	Standard
	Mn	55	28090211.5	1.2	**3736.5671**	85.696	2.3	ug/L	747	Standard
	Co	59	552502.8	1.4	**77.4533**	1.929	2.5	ug/L	234	Standard
	Ni	60	270132.5	1.6	**133.5117**	3.898	2.9	ug/L	215	Standard
	Cu	65	201432.2	2.2	**100.1412**	3.335	3.3	ug/L	104	Standard
	Zn	66	320182.0	2.8	**271.0655**	11.001	4.1	ug/L	363	Standard
>	Ge	72	534719.5	1.4				ug/L	495142	Standard
	As	75	41934.7	2.3	**34.7370**	1.197	3.4	ug/L	-50	Standard
	Se	82	2480.2	1.5	**20.7934**	0.607	2.9	ug/L	14	Standard
	Se-1	77	1799.1	2.3	**22.1951**	0.836	3.8	ug/L	42	Standard
>	Ga	71	63861.1	4.0				mg/L	42	Standard
	Rb	85	554637.9	1.8				ug/L	17	Standard
	Y	89	1250941.0	3.4				ug/L	417098	Standard
>	Rh	103	88.3	8.6				ug/L	13	Standard
	Mo	98	1454.1	4.8	**0.5217**	0.033	6.4	ug/L	11	Standard
	Ag	107	4814.4	4.1	**0.7006**	0.018	2.6	ug/L	68	Standard
	Cd	111	56224.0	1.3	**25.9728**	0.111	0.4	mg/L	2	Standard
	Cd	114	138439.7	3.6	**26.0717**	0.493	1.9	ug/L	12	Standard
>	In	115	608045.3	1.7				ug/L	587441	Standard
	Sn	118	518.3	2.0	**-0.0090**	0.001	16.4	ug/L	497	Standard
	Sb	123	2325.7	1.7	**0.4771**	0.008	1.7	ug/L	78	Standard
	Ba	135	587827.7	1.5	**277.4602**	0.795	0.3	ug/L	24	Standard
	Ce	140	2927642.8	0.6				ug/L	25	Standard
>	Tb	159	1033681.4	2.7				ug/L	939701	Standard
	Ho	165	68191.4	0.6				ug/L	10	Standard
	Tl	203	212892.3	0.8	**25.4932**	0.083	0.3	ug/L	4	Standard
	Tl	205	191048.7	1.2	**26.3598**	0.091	0.3	ug/L	2	Standard
	Pb	206	260602.6	2.5	**48.0843**	0.971	2.0	ug/L	240	Standard
	Pb	207	216580.3	1.8	**44.9759**	0.652	1.4	ug/L	198	Standard
	Pb	208	918219.8	1.2	**48.2652**	0.290	0.6	ug/L	883	Standard
	U	238	196774.7	2.4	**28.2131**	0.460	1.6	ug/L	3	Standard
>	Bi	209	530811.0	1.1				ug/L	523984	Standard

Sample ID: L1606088803S WG573099-04
Report Date/Time: Wednesday, June 29, 2016 15:22:23
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**12.2205**	21.158	173.1	mg/L	0	Standard
Mg	24	365.0	3.6	**0.2314**	0.007	3.2	mg/L	37	Standard
K	39	1516.7	5.6	**1.9963**	0.137	6.9	mg/L	38	Standard
Ca	43	726.7	5.8	**26.3770**	1.992	7.6	mg/L	10	Standard
Fe	54	108682.6	3.2	**57.8341**	0.799	1.4	mg/L	179	Standard
Fe	57	27718.7	1.4	**58.2169**	0.892	1.5	mg/L	335	Standard
Sc-1	45	96992.9	2.3				mg/L	61029	Standard
Cl	35	26216.7	4.4				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	4344.0	8.1				ug/L	1547	Standard
P	31	22273.0	3.2				ug/L	15286	Standard
S	34	2523.5	5.1				ug/L	1572	Standard
Sr	88	263.3	11.4				ug/L	73	Standard
C	12	56.7	10.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	52.9				mg/L	3	Standard
Dy	164	111985.6	3.7				mg/L	6	Standard
Ho-1	165	68191.4	0.6				mg/L	10	Standard
Er	166	57569.3	4.1				mg/L	13	Standard
I	127	6513.1	8.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		118.857	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		107.993	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606088803S WG573099-04
Report Date/Time: Wednesday, June 29, 2016 15:22:23
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.507
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.303
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ni 60 Upper, S, EEE	Ni	60	

Sample ID: L1606088803S WG573099-04
Report Date/Time: Wednesday, June 29, 2016 15:22:23
Page 3

Approved: July 04, 2016

Cu 65 Upper, S, EEE	Cu	65
Zn 66 Upper, S, EEE	Zn	66
Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088803SD WG573099-05
Sample Date/Time: Wednesday, June 29, 2016 15:23:18
Number of Replicates: 3
Autosampler Position: 340
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86779.6	2.9				ug/L	74699	Standard
	Be	9	26783.7	4.5	**26.1889**	0.607	2.3	ug/L	8	Standard
	Al	27	17495490.1	2.1	**102.4852**	1.702	1.7	ug/L	1493	Standard
	Sc	45	84971.1	3.4				ug/L	61029	Standard
	Ti	47	11032.7	2.5	**25.6633**	0.769	3.0	ug/L	35	Standard
	V	51	492228.0	0.7	**57.5050**	0.764	1.3	ug/L	1435	Standard
	Cr	52	556079.4	1.4	**65.4064**	1.154	1.8	ug/L	7635	Standard
	Cr	53	67044.7	1.2	**63.8916**	0.827	1.3	ug/L	363	Standard
	Mn	55	16288704.5	1.7	**2071.2097**	47.470	2.3	ug/L	747	Standard
	Co	59	487642.9	2.1	**65.3439**	1.764	2.7	ug/L	234	Standard
	Ni	60	230286.3	2.0	**108.7653**	2.155	2.0	ug/L	215	Standard
	Cu	65	186555.0	1.4	**88.6352**	1.512	1.7	ug/L	104	Standard
	Zn	66	269449.8	1.5	**217.9309**	3.745	1.7	ug/L	363	Standard
>	**Ge**	72	559301.6	0.7				ug/L	495142	Standard
	As	75	32569.4	2.0	**25.8033**	0.648	2.5	ug/L	-50	Standard
	Se	82	2764.3	2.2	**22.1623**	0.630	2.8	ug/L	14	Standard
	Se-1	77	1950.8	4.1	**23.0256**	1.042	4.5	ug/L	42	Standard
>	**Ga**	71	55862.8	2.2				mg/L	42	Standard
	Rb	85	472783.4	2.1				ug/L	17	Standard
	Y	89	931018.8	1.5				ug/L	417098	Standard
>	**Rh**	103	36.7	47.9				ug/L	13	Standard
	Mo	98	833.5	3.5	**0.3003**	0.015	5.0	ug/L	11	Standard
	Ag	107	3972.2	0.3	**0.5833**	0.011	1.9	ug/L	68	Standard
	Cd	111	54996.2	0.6	**25.7204**	0.316	1.2	mg/L	2	Standard
	Cd	114	133681.6	2.4	**25.4888**	0.206	0.8	ug/L	12	Standard
>	**In**	115	600647.8	1.6				ug/L	587441	Standard
	Sn	118	466.7	16.3	**-0.0167**	0.012	70.0	ug/L	497	Standard
	Sb	123	3387.1	2.7	**0.7053**	0.008	1.2	ug/L	78	Standard
	Ba	135	467382.4	1.0	**223.3343**	2.175	1.0	ug/L	24	Standard
	Ce	140	1516729.6	1.2				ug/L	25	Standard
>	**Tb**	159	1007426.2	2.1				ug/L	939701	Standard
	Ho	165	39640.1	1.9				ug/L	10	Standard
	Tl	203	211811.1	1.3	**24.5598**	0.357	1.5	ug/L	4	Standard
	Tl	205	187359.9	1.8	**25.0339**	0.597	2.4	ug/L	2	Standard
	Pb	206	219015.7	1.1	**39.1225**	0.790	2.0	ug/L	240	Standard
	Pb	207	182856.8	0.5	**36.7608**	0.679	1.8	ug/L	198	Standard
	Pb	208	768036.7	1.8	**39.0813**	0.960	2.5	ug/L	883	Standard
	U	238	187429.9	1.2	**26.0258**	0.546	2.1	ug/L	3	Standard
>	**Bi**	209	548237.3	1.8				ug/L	523984	Standard

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 1

Approved: July 04, 2016

Na	23	6.7	43.3	**55.5842**	25.460	45.8	mg/L	0	Standard
Mg	24	348.3	15.5	**0.2571**	0.039	15.2	mg/L	37	Standard
K	39	1421.7	3.5	**2.1441**	0.098	4.5	mg/L	38	Standard
Ca	43	383.3	17.4	**14.9477**	3.560	23.8	mg/L	10	Standard
Fe	54	86673.1	2.4	**52.6645**	1.510	2.9	mg/L	179	Standard
Fe	57	21795.7	3.1	**52.2085**	3.369	6.5	mg/L	335	Standard
> **Sc-1**	45	84971.1	3.4				mg/L	61029	Standard
Cl	35	24892.4	3.3				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	3257.0	4.5				ug/L	1547	Standard
P	31	20836.0	2.5				ug/L	15286	Standard
S	34	2406.9	4.4				ug/L	1572	Standard
Sr	88	191.7	46.0				ug/L	73	Standard
C	12	76.7	15.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	20.0				mg/L	3	Standard
Dy	164	63526.1	1.7				mg/L	6	Standard
Ho-1	165	39640.1	1.9				mg/L	10	Standard
Er	166	35224.2	2.0				mg/L	13	Standard
I	127	5139.2	9.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		116.173	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		112.958	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.248
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.629
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ni 60 Upper, S, EEE	Ni	60	

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 3

Approved: July 04, 2016

Zn 66 Upper, S, EEE	Zn	66
Ba 135 Upper, S, EEE	Ba	135

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088301
Sample Date/Time: Wednesday, June 29, 2016 15:26:29
Number of Replicates: 3
Autosampler Position: 341
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84384.4	1.7				ug/L	74699	Standard
	Be	9	1073.4	1.5	**1.0872**	0.032	3.0	ug/L	8	Standard
	Al	27	11940771.9	2.8	**71.9363**	2.792	3.9	ug/L	1493	Standard
	Sc	45	76626.6	1.2				ug/L	61029	Standard
	Ti	47	4374.3	2.0	**10.2834**	0.247	2.4	ug/L	35	Standard
	V	51	162703.8	1.0	**19.2014**	0.098	0.5	ug/L	1435	Standard
	Cr	52	132016.8	1.7	**15.0167**	0.347	2.3	ug/L	7635	Standard
	Cr	53	16153.6	0.8	**15.4073**	0.063	0.4	ug/L	363	Standard
	Mn	55	3262876.5	0.6	**421.8414**	5.005	1.2	ug/L	747	Standard
	Co	59	67832.8	0.7	**9.2120**	0.089	1.0	ug/L	234	Standard
	Ni	60	42623.2	1.3	**20.3743**	0.274	1.3	ug/L	215	Standard
	Cu	65	31205.8	0.3	**14.9935**	0.149	1.0	ug/L	104	Standard
	Zn	66	53134.7	1.9	**43.3429**	0.711	1.6	ug/L	363	Standard
>	**Ge**	72	549915.9	0.8				ug/L	495142	Standard
	As	75	3335.6	3.0	**2.7409**	0.080	2.9	ug/L	-50	Standard
	Se	82	46.6	19.0	**0.2449**	0.076	30.9	ug/L	14	Standard
	Se-1	77	111.7	5.1	**0.7542**	0.066	8.8	ug/L	42	Standard
>	**Ga**	71	17852.2	2.1				mg/L	42	Standard
	Rb	85	173572.1	2.0				ug/L	17	Standard
	Y	89	703667.4	2.2				ug/L	417098	Standard
>	**Rh**	103	33.3	34.6				ug/L	13	Standard
	Mo	98	357.4	3.8	**0.1158**	0.007	6.2	ug/L	11	Standard
	Ag	107	550.0	6.6	**0.0636**	0.007	11.1	ug/L	68	Standard
	Cd	111	273.3	3.2	**0.1105**	0.006	5.2	mg/L	2	Standard
	Cd	114	617.1	4.0	**0.1069**	0.006	5.7	ug/L	12	Standard
>	**In**	115	648944.8	3.3				ug/L	587441	Standard
	Sn	118	606.7	6.6	**-0.0008**	0.003	418.3	ug/L	497	Standard
	Sb	123	137.4	14.1	**0.0224**	0.003	13.8	ug/L	78	Standard
	Ba	135	318337.1	2.2	**140.8176**	2.386	1.7	ug/L	24	Standard
	Ce	140	1211648.5	1.9				ug/L	25	Standard
>	**Tb**	159	1041543.0	2.5				ug/L	939701	Standard
	Ho	165	19666.1	3.2				ug/L	10	Standard
	Tl	203	1035.0	3.0	**0.1143**	0.004	3.7	ug/L	4	Standard
	Tl	205	953.4	8.0	**0.1233**	0.009	7.5	ug/L	2	Standard
	Pb	206	127279.7	1.1	**21.9443**	0.442	2.0	ug/L	240	Standard
	Pb	207	103046.6	1.5	**19.9907**	0.426	2.1	ug/L	198	Standard
	Pb	208	441148.5	1.4	**21.6612**	0.196	0.9	ug/L	883	Standard
	U	238	5034.2	0.5	**0.6713**	0.005	0.8	ug/L	3	Standard
>	**Bi**	209	567168.2	1.1				ug/L	523984	Standard

Sample ID: L1606088301
Report Date/Time: Wednesday, June 29, 2016 15:28:46
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	340.0	3.9	**0.2844**	0.011	3.7	mg/L	37	Standard
K	39	688.3	2.3	**1.0975**	0.019	1.8	mg/L	38	Standard
Ca	43	145.0	15.8	**4.7718**	1.160	24.3	mg/L	10	Standard
Fe	54	17324.9	3.4	**11.5598**	0.337	2.9	mg/L	179	Standard
Fe	57	4550.7	3.6	**11.2444**	0.399	3.5	mg/L	335	Standard
Sc-1	45	76626.6	1.2				mg/L	61029	Standard
Cl	35	25558.2	1.3				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	2263.5	8.1				ug/L	1547	Standard
P	31	20323.6	0.9				ug/L	15286	Standard
S	34	2258.5	1.1				ug/L	1572	Standard
Sr	88	101.7	10.2				ug/L	73	Standard
C	12	116.7	13.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	49.5				mg/L	3	Standard
Dy	164	32996.6	2.0				mg/L	6	Standard
Ho-1	165	19666.1	3.2				mg/L	10	Standard
Er	166	17349.9	3.1				mg/L	13	Standard
I	127	11507.7	1.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		112.966	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		111.062	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606088301
Report Date/Time: Wednesday, June 29, 2016 15:28:46
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	110.470
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.241
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606088301
Report Date/Time: Wednesday, June 29, 2016 15:28:46
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201
Sample Date/Time: Wednesday, June 29, 2016 15:29:41
Number of Replicates: 3
Autosampler Position: 342
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84277.1	1.0				ug/L	74699	Standard
	Be	9	1593.4	6.0	**1.6114**	0.083	5.1	ug/L	8	Standard
	Al	27	15243679.1	1.9	**91.9299**	1.967	2.1	ug/L	1493	Standard
	Sc	45	83538.1	3.1				ug/L	61029	Standard
	Ti	47	10307.8	1.1	**23.4331**	0.360	1.5	ug/L	35	Standard
	V	51	410563.1	0.8	**46.8597**	0.438	0.9	ug/L	1435	Standard
	Cr	52	261972.4	0.3	**29.5747**	0.180	0.6	ug/L	7635	Standard
	Cr	53	32057.2	1.7	**29.6933**	0.417	1.4	ug/L	363	Standard
	Mn	55	12908980.1	3.2	**1604.9549**	58.354	3.6	ug/L	747	Standard
	Co	59	181241.6	3.1	**23.7226**	0.846	3.6	ug/L	234	Standard
	Ni	60	66685.8	0.9	**30.7063**	0.259	0.8	ug/L	215	Standard
	Cu	65	73374.2	0.7	**34.0220**	0.398	1.2	ug/L	104	Standard
	Zn	66	143517.2	0.9	**113.2755**	1.322	1.2	ug/L	363	Standard
>	**Ge**	72	572023.7	0.5				ug/L	495142	Standard
	As	75	16668.5	2.4	**12.9416**	0.353	2.7	ug/L	-50	Standard
	Se	82	187.7	6.4	**1.3421**	0.095	7.1	ug/L	14	Standard
	Se-1	77	232.0	7.5	**2.1278**	0.217	10.2	ug/L	42	Standard
>	**Ga**	71	37605.0	3.4				mg/L	42	Standard
	Rb	85	431124.7	1.0				ug/L	17	Standard
	Y	89	907846.1	0.7				ug/L	417098	Standard
>	**Rh**	103	28.3	27.0				ug/L	13	Standard
	Mo	98	1317.8	2.4	**0.4515**	0.009	1.9	ug/L	11	Standard
	Ag	107	1338.4	1.8	**0.1770**	0.004	2.5	ug/L	68	Standard
	Cd	111	803.3	6.8	**0.3473**	0.022	6.2	mg/L	2	Standard
	Cd	114	1888.8	7.2	**0.3384**	0.023	6.7	ug/L	12	Standard
>	**In**	115	635240.6	1.1				ug/L	587441	Standard
	Sn	118	608.3	2.5	**0.0016**	0.002	109.2	ug/L	497	Standard
	Sb	123	242.0	3.6	**0.0438**	0.002	4.9	ug/L	78	Standard
	Ba	135	645480.7	0.7	**291.6323**	1.817	0.6	ug/L	24	Standard
	Ce	140	2459220.7	0.8				ug/L	25	Standard
>	**Tb**	159	1043691.8	1.6				ug/L	939701	Standard
	Ho	165	34592.8	2.2				ug/L	10	Standard
	Tl	203	3164.7	1.5	**0.3540**	0.007	2.0	ug/L	4	Standard
	Tl	205	2843.6	5.4	**0.3684**	0.020	5.4	ug/L	2	Standard
	Pb	206	325793.2	0.7	**56.4391**	0.926	1.6	ug/L	240	Standard
	Pb	207	262124.0	0.3	**51.1005**	0.344	0.7	ug/L	198	Standard
	Pb	208	1122453.0	1.2	**55.3927**	1.161	2.1	ug/L	883	Standard
	U	238	18503.6	0.6	**2.4863**	0.037	1.5	ug/L	3	Standard
>	**Bi**	209	565563.6	0.9				ug/L	523984	Standard

Sample ID: L1606091201
Report Date/Time: Wednesday, June 29, 2016 15:31:58
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	148.3	19.2	**0.0751**	0.029	38.8	mg/L	37	Standard
K	39	1071.7	2.8	**1.6167**	0.038	2.4	mg/L	38	Standard
Ca	43	301.7	16.4	**11.4520**	2.661	23.2	mg/L	10	Standard
Fe	54	45977.7	1.1	**28.3556**	0.839	3.0	mg/L	179	Standard
Fe	57	11742.9	4.0	**28.0839**	0.746	2.7	mg/L	335	Standard
Sc-1	45	83538.1	3.1				mg/L	61029	Standard
Cl	35	25586.9	0.8				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	3940.5	1.8				ug/L	1547	Standard
P	31	20170.1	2.1				ug/L	15286	Standard
S	34	2405.2	3.2				ug/L	1572	Standard
Sr	88	111.7	6.8				ug/L	73	Standard
C	12	316.7	15.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	146.7	35.0				mg/L	3	Standard
Dy	164	57194.9	2.2				mg/L	6	Standard
Ho-1	165	34592.8	2.2				mg/L	10	Standard
Er	166	32892.3	1.1				mg/L	13	Standard
I	127	48256.8	4.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		112.823	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.527	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606091201
Report Date/Time: Wednesday, June 29, 2016 15:31:58
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	108.137
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	107.935
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606091201
Report Date/Time: Wednesday, June 29, 2016 15:31:58
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201PS WG573717-03
Sample Date/Time: Wednesday, June 29, 2016 15:32:52
Number of Replicates: 3
Autosampler Position: 343
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	84925.7	1.3				ug/L	74699	Standard
	Be	9	54409.2	0.5	**54.3741**	0.919	1.7	ug/L	8	Standard
	Al	27	15619580.7	1.4	**93.4869**	2.147	2.3	ug/L	1493	Standard
	Sc	45	84914.0	2.3				ug/L	61029	Standard
	Ti	47	10489.6	2.4	**24.7035**	0.475	1.9	ug/L	35	Standard
	V	51	847646.1	0.7	**100.4344**	1.074	1.1	ug/L	1435	Standard
	Cr	52	696222.3	1.3	**83.2031**	0.580	0.7	ug/L	7635	Standard
	Cr	53	86425.9	2.6	**83.5032**	1.546	1.9	ug/L	363	Standard
	Mn	55	12964778.0	1.1	**1669.4256**	16.996	1.0	ug/L	747	Standard
	Co	59	568224.0	0.4	**77.1151**	0.744	1.0	ug/L	234	Standard
	Ni	60	176707.0	1.9	**84.4902**	0.918	1.1	ug/L	215	Standard
	Cu	65	181984.3	1.8	**87.5571**	0.864	1.0	ug/L	104	Standard
	Zn	66	208968.8	1.8	**171.0579**	1.957	1.1	ug/L	363	Standard
>	Ge	72	552256.5	0.8				ug/L	495142	Standard
	As	75	82371.4	1.6	**65.9929**	0.953	1.4	ug/L	-50	Standard
	Se	82	6597.3	1.6	**53.7542**	0.590	1.1	ug/L	14	Standard
	Se-1	77	4584.4	1.6	**55.6563**	1.034	1.9	ug/L	42	Standard
>	Ga	71	37593.3	1.5				mg/L	42	Standard
	Rb	85	428376.8	2.4				ug/L	17	Standard
	Y	89	890825.8	0.7				ug/L	417098	Standard
>	Rh	103	50.0	10.0				ug/L	13	Standard
	Mo	98	1335.6	2.9	**0.4503**	0.014	3.2	ug/L	11	Standard
	Ag	107	364007.1	1.3	**50.8126**	0.330	0.6	ug/L	68	Standard
	Cd	111	115695.5	1.2	**50.3468**	0.384	0.8	mg/L	2	Standard
	Cd	114	280904.1	1.9	**49.8365**	0.143	0.3	ug/L	12	Standard
>	In	115	645589.3	1.8				ug/L	587441	Standard
	Sn	118	755.0	4.3	**0.0229**	0.003	13.2	ug/L	497	Standard
	Sb	123	254065.1	1.3	**49.5173**	0.623	1.3	ug/L	78	Standard
	Ba	135	752068.0	1.2	**334.3566**	2.055	0.6	ug/L	24	Standard
	Ce	140	2468196.2	1.3				ug/L	25	Standard
>	Tb	159	1069312.5	1.3				ug/L	939701	Standard
	Ho	165	35102.3	1.1				ug/L	10	Standard
	Tl	203	458369.6	1.1	**50.4504**	0.349	0.7	ug/L	4	Standard
	Tl	205	409805.1	0.9	**51.9806**	1.259	2.4	ug/L	2	Standard
	Pb	206	619020.9	0.8	**105.0729**	1.076	1.0	ug/L	240	Standard
	Pb	207	511972.0	0.6	**97.8107**	1.381	1.4	ug/L	198	Standard
	Pb	208	2150459.1	0.4	**103.9876**	1.211	1.2	ug/L	883	Standard
	U	238	395808.0	0.4	**52.1735**	0.789	1.5	ug/L	3	Standard
>	Bi	209	577539.6	1.5				ug/L	523984	Standard

Sample ID: L1606091201PS WG573717-03
Report Date/Time: Wednesday, June 29, 2016 15:35:09
Page 1

Approved: July 04, 2016

Na	23	6.7	43.3	**54.9152**	22.487	40.9	mg/L	0	Standard
Mg	24	118.3	17.1	**0.0444**	0.016	36.3	mg/L	37	Standard
K	39	1026.7	7.6	**1.5167**	0.116	7.6	mg/L	38	Standard
Ca	43	296.7	14.3	**10.9789**	2.237	20.4	mg/L	10	Standard
Fe	54	45524.1	0.4	**27.6104**	0.565	2.0	mg/L	179	Standard
Fe	57	12291.7	1.9	**28.9717**	1.259	4.3	mg/L	335	Standard
> **Sc-1**	45	84914.0	2.3				mg/L	61029	Standard
Cl	35	25819.3	0.8				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	3963.9	6.7				ug/L	1547	Standard
P	31	20629.0	3.4				ug/L	15286	Standard
S	34	2280.2	5.7				ug/L	1572	Standard
Sr	88	103.3	12.2				ug/L	73	Standard
C	12	333.3	17.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	140.0	7.1				mg/L	3	Standard
Dy	164	56592.8	3.3				mg/L	6	Standard
Ho-1	165	35102.3	1.1				mg/L	10	Standard
Er	166	32397.9	1.1				mg/L	13	Standard
I	127	48815.3	4.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.691	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.535	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091201PS WG573717-03
Report Date/Time: Wednesday, June 29, 2016 15:35:09
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.899
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.221
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
V 51 Upper, S, EEE	V	51	
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606091201PS WG573717-03
Report Date/Time: Wednesday, June 29, 2016 15:35:09
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135
Pb 206 Upper, S, EEE	Pb	206
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201SDL WG573717-04
Sample Date/Time: Wednesday, June 29, 2016 15:36:04
Number of Replicates: 3
Autosampler Position: 344
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83450.9	2.1				ug/L	74699	Standard
	Be	9	378.3	16.7	**0.3916**	0.058	14.9	ug/L	8	Standard
	Al	27	3082760.3	2.5	**18.7625**	0.099	0.5	ug/L	1493	Standard
	Sc	45	72303.8	2.1				ug/L	61029	Standard
	Ti	47	2033.1	4.6	**4.7237**	0.243	5.1	ug/L	35	Standard
	V	51	79797.1	0.9	**9.3216**	0.052	0.6	ug/L	1435	Standard
	Cr	52	57744.9	2.6	**5.9945**	0.158	2.6	ug/L	7635	Standard
	Cr	53	6483.1	2.8	**5.9884**	0.155	2.6	ug/L	363	Standard
	Mn	55	2512242.8	1.3	**324.8710**	1.825	0.6	ug/L	747	Standard
	Co	59	34712.7	2.2	**4.6975**	0.071	1.5	ug/L	234	Standard
	Ni	60	13460.0	2.9	**6.3513**	0.158	2.5	ug/L	215	Standard
	Cu	65	15009.4	2.6	**7.1598**	0.141	2.0	ug/L	104	Standard
	Zn	66	29256.2	2.4	**23.6692**	0.389	1.6	ug/L	363	Standard
>	**Ge**	72	549686.2	0.8				ug/L	495142	Standard
	As	75	3156.6	2.0	**2.5979**	0.033	1.3	ug/L	-50	Standard
	Se	82	47.1	17.6	**0.2490**	0.070	28.2	ug/L	14	Standard
	Se-1	77	77.0	11.9	**0.3268**	0.106	32.4	ug/L	42	Standard
>	**Ga**	71	7136.7	1.1				mg/L	42	Standard
	Rb	85	84221.8	1.7				ug/L	17	Standard
	Y	89	569587.4	0.8				ug/L	417098	Standard
>	**Rh**	103	25.0	91.7				ug/L	13	Standard
	Mo	98	271.1	6.9	**0.0847**	0.006	6.6	ug/L	11	Standard
	Ag	107	378.0	10.8	**0.0387**	0.005	13.9	ug/L	68	Standard
	Cd	111	185.8	10.3	**0.0712**	0.009	12.3	mg/L	2	Standard
	Cd	114	406.5	6.5	**0.0684**	0.004	6.4	ug/L	12	Standard
>	**In**	115	660230.6	1.0				ug/L	587441	Standard
	Sn	118	501.7	8.5	**-0.0183**	0.006	32.3	ug/L	497	Standard
	Sb	123	1797.9	20.8	**0.3384**	0.070	20.8	ug/L	78	Standard
	Ba	135	129725.4	1.6	**56.3802**	0.972	1.7	ug/L	24	Standard
	Ce	140	480517.5	2.7				ug/L	25	Standard
>	**Tb**	159	1042248.3	0.8				ug/L	939701	Standard
	Ho	165	6823.2	4.6				ug/L	10	Standard
	Tl	203	727.7	7.7	**0.0794**	0.006	7.1	ug/L	4	Standard
	Tl	205	616.7	4.2	**0.0795**	0.002	2.4	ug/L	2	Standard
	Pb	206	64359.4	2.2	**11.0114**	0.128	1.2	ug/L	240	Standard
	Pb	207	51665.3	2.6	**9.9429**	0.134	1.3	ug/L	198	Standard
	Pb	208	222946.4	0.6	**10.8660**	0.247	2.3	ug/L	883	Standard
	U	238	3701.1	2.3	**0.4902**	0.006	1.1	ug/L	3	Standard
>	**Bi**	209	569638.6	2.0				ug/L	523984	Standard

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:38:21
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	61.7	28.5	**0.0019**	0.018	942.5	mg/L	37	Standard
K	39	203.3	18.5	**0.2649**	0.067	25.4	mg/L	38	Standard
Ca	43	75.0	26.7	**1.4632**	0.987	67.5	mg/L	10	Standard
Fe	54	9325.9	3.5	**6.5358**	0.204	3.1	mg/L	179	Standard
Fe	57	2630.2	2.1	**6.4718**	0.055	0.8	mg/L	335	Standard
> **Sc-1**	45	72303.8	2.1				mg/L	61029	Standard
Cl	35	24972.6	0.9				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2286.8	8.4				ug/L	1547	Standard
P	31	19043.6	4.4				ug/L	15286	Standard
S	34	2180.2	1.2				ug/L	1572	Standard
Sr	88	81.7	35.9				ug/L	73	Standard
C	12	93.3	22.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	57.3				mg/L	3	Standard
Dy	164	11677.0	2.4				mg/L	6	Standard
Ho-1	165	6823.2	4.6				mg/L	10	Standard
Er	166	6424.7	2.9				mg/L	13	Standard
I	127	12089.8	7.2				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		111.717	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		111.016	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:38:21
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.391
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.713
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201SDL WG573717-04
Sample Date/Time: Wednesday, June 29, 2016 15:39:15
Number of Replicates: 3
Autosampler Position: 345
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80082.8	1.5				ug/L	74699	Standard
	Be	9	80.0	10.8	**0.0924**	0.009	10.2	ug/L	8	Standard
	Al	27	638793.5	0.9	**4.0416**	0.057	1.4	ug/L	1493	Standard
	Sc	45	66977.8	2.5				ug/L	61029	Standard
	Ti	47	422.3	5.3	**0.8899**	0.044	4.9	ug/L	35	Standard
	V	51	17807.2	0.8	**1.9190**	0.038	2.0	ug/L	1435	Standard
	Cr	52	18051.7	1.5	**1.1605**	0.015	1.3	ug/L	7635	Standard
	Cr	53	1846.8	4.0	**1.4619**	0.067	4.6	ug/L	363	Standard
	Mn	55	540671.4	1.9	**69.4959**	1.549	2.2	ug/L	747	Standard
	Co	59	7520.2	3.3	**0.9841**	0.038	3.8	ug/L	234	Standard
	Ni	60	2912.3	1.6	**1.2710**	0.019	1.5	ug/L	215	Standard
	Cu	65	3203.0	1.2	**1.4395**	0.003	0.2	ug/L	104	Standard
	Zn	66	8550.1	1.3	**6.5648**	0.083	1.3	ug/L	363	Standard
>	**Ge**	72	552153.4	1.0				ug/L	495142	Standard
	As	75	624.0	5.3	**0.5591**	0.024	4.4	ug/L	-50	Standard
	Se	82	21.8	25.1	**0.0396**	0.043	108.7	ug/L	14	Standard
	Se-1	77	57.0	8.0	**0.0772**	0.050	64.2	ug/L	42	Standard
>	**Ga**	71	1470.1	2.9				mg/L	42	Standard
	Rb	85	17425.0	3.8				ug/L	17	Standard
	Y	89	497203.9	1.3				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	65.7	10.1	**0.0165**	0.003	15.8	ug/L	11	Standard
	Ag	107	156.7	8.7	**0.0087**	0.002	18.0	ug/L	68	Standard
	Cd	111	38.9	11.8	**0.0089**	0.002	23.3	mg/L	2	Standard
	Cd	114	108.8	14.4	**0.0170**	0.003	15.1	ug/L	12	Standard
>	**In**	115	651862.8	1.7				ug/L	587441	Standard
	Sn	118	480.0	4.2	**-0.0206**	0.003	16.7	ug/L	497	Standard
	Sb	123	490.3	31.4	**0.0908**	0.031	34.3	ug/L	78	Standard
	Ba	135	26415.7	0.1	**11.6170**	0.204	1.8	ug/L	24	Standard
	Ce	140	98329.7	0.4				ug/L	25	Standard
>	**Tb**	159	1007324.2	1.5				ug/L	939701	Standard
	Ho	165	1405.1	5.7				ug/L	10	Standard
	Tl	203	169.0	1.6	**0.0177**	0.000	0.8	ug/L	4	Standard
	Tl	205	175.0	8.6	**0.0234**	0.002	8.3	ug/L	2	Standard
	Pb	206	13505.0	1.4	**2.3274**	0.057	2.5	ug/L	240	Standard
	Pb	207	10900.6	1.3	**2.1089**	0.044	2.1	ug/L	198	Standard
	Pb	208	47029.6	1.3	**2.3051**	0.014	0.6	ug/L	883	Standard
	U	238	743.4	3.4	**0.0980**	0.002	2.1	ug/L	3	Standard
>	**Bi**	209	552389.3	1.4				ug/L	523984	Standard

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:41:32
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	33.3	17.3	**-0.0260**	0.006	22.5	mg/L	37	Standard
K	39	50.0	36.1	**-0.0133**	0.037	281.6	mg/L	38	Standard
Ca	43	41.7	48.5	**-0.1354**	1.146	846.9	mg/L	10	Standard
Fe	54	2020.3	1.1	**1.4233**	0.051	3.6	mg/L	179	Standard
Fe	57	728.4	10.0	**1.1787**	0.168	14.2	mg/L	335	Standard
> **Sc-1**	45	66977.8	2.5				mg/L	61029	Standard
Cl	35	24720.8	1.1				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	1860.1	4.8				ug/L	1547	Standard
P	31	18349.4	1.0				ug/L	15286	Standard
S	34	1845.1	4.4				ug/L	1572	Standard
Sr	88	65.0	27.7				ug/L	73	Standard
C	12	66.7	31.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	2304.6	5.8				mg/L	6	Standard
Ho-1	165	1405.1	5.7				mg/L	10	Standard
Er	166	1370.1	4.1				mg/L	13	Standard
I	127	4489.0	15.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		107.208	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.514	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:41:32
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.966
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.421
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 15:42:28
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79858.3	1.9				ug/L	74699	Standard
	Be	9	50198.1	2.5	**53.3484**	1.401	2.6	ug/L	8	Standard
	Al	27	8066966.3	0.3	**51.3427**	0.865	1.7	ug/L	1493	Standard
	Sc	45	68357.2	2.3				ug/L	61029	Standard
	Ti	47	43587.6	0.3	**105.7066**	1.724	1.6	ug/L	35	Standard
	V	51	433103.8	0.8	**52.5653**	0.576	1.1	ug/L	1435	Standard
	Cr	52	431112.3	0.9	**52.4990**	0.789	1.5	ug/L	7635	Standard
	Cr	53	52653.0	1.0	**52.0862**	0.689	1.3	ug/L	363	Standard
	Mn	55	405815.5	1.7	**53.4827**	1.214	2.3	ug/L	747	Standard
	Co	59	390955.7	1.0	**54.4418**	1.354	2.5	ug/L	234	Standard
	Ni	60	105213.1	1.7	**51.5742**	0.488	0.9	ug/L	215	Standard
	Cu	65	104377.6	0.2	**51.4980**	0.905	1.8	ug/L	104	Standard
	Zn	66	61654.9	0.7	**51.4853**	1.289	2.5	ug/L	363	Standard
>	**Ge**	72	538231.7	1.8				ug/L	495142	Standard
	As	75	61802.6	0.2	**50.8282**	0.881	1.7	ug/L	-50	Standard
	Se	82	6211.6	1.2	**51.9313**	0.364	0.7	ug/L	14	Standard
	Se-1	77	4323.6	2.0	**53.8508**	1.659	3.1	ug/L	42	Standard
>	**Ga**	71	60.0	28.9				mg/L	42	Standard
	Rb	85	955.0	5.2				ug/L	17	Standard
	Y	89	475770.1	1.3				ug/L	417098	Standard
>	**Rh**	103	25.0	34.6				ug/L	13	Standard
	Mo	98	290643.5	0.8	**97.6712**	0.653	0.7	ug/L	11	Standard
	Ag	107	354998.2	1.8	**48.7852**	1.288	2.6	ug/L	68	Standard
	Cd	111	115105.3	1.2	**49.3038**	0.151	0.3	mg/L	2	Standard
	Cd	114	278592.4	1.8	**48.6531**	0.360	0.7	ug/L	12	Standard
>	**In**	115	655825.0	1.0				ug/L	587441	Standard
	Sn	118	319548.2	1.2	**48.9035**	0.256	0.5	ug/L	497	Standard
	Sb	123	249456.0	1.2	**47.8604**	0.919	1.9	ug/L	78	Standard
	Ba	135	108215.4	0.8	**47.3455**	0.546	1.2	ug/L	24	Standard
	Ce	140	256.7	76.3				ug/L	25	Standard
>	**Tb**	159	1025574.2	1.9				ug/L	939701	Standard
	Ho	165	15.0	0.0				ug/L	10	Standard
	Tl	203	443897.9	1.2	**50.1327**	1.058	2.1	ug/L	4	Standard
	Tl	205	398087.0	0.7	**51.8003**	0.814	1.6	ug/L	2	Standard
	Pb	206	280554.3	0.5	**48.8226**	0.728	1.5	ug/L	240	Standard
	Pb	207	249585.9	0.6	**48.8846**	0.533	1.1	ug/L	198	Standard
	Pb	208	1026713.8	0.6	**50.9050**	1.131	2.2	ug/L	883	Standard
	U	238	372550.1	0.8	**50.3866**	1.127	2.2	ug/L	3	Standard
>	**Bi**	209	562933.7	1.7				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:44:45
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4292.3	1.6	**4.8821**	0.035	0.7	mg/L	37	Standard
K	39	2345.2	3.1	**4.5133**	0.080	1.8	mg/L	38	Standard
Ca	43	103.3	10.1	**3.3018**	0.528	16.0	mg/L	10	Standard
Fe	54	6681.2	2.2	**4.9206**	0.153	3.1	mg/L	179	Standard
Fe	57	1928.5	4.0	**4.7824**	0.344	7.2	mg/L	335	Standard
Sc-1	45	68357.2	2.3				mg/L	61029	Standard
Cl	35	26753.6	0.7				ug/L	20195	Standard
Kr	83	3.3	45.8				ug/L	4	Standard
Br	81	1833.4	8.1				ug/L	1547	Standard
P	31	20395.4	2.2				ug/L	15286	Standard
S	34	2408.5	8.3				ug/L	1572	Standard
Sr	88	85.0	25.6				ug/L	73	Standard
C	12	53.3	28.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.2	35.5				mg/L	6	Standard
Ho-1	165	15.0	0.0				mg/L	10	Standard
Er	166	30.0	57.7				mg/L	13	Standard
I	127	2536.9	16.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	106.697		
Al	27	102.685		
Sc	45			
Ti	47	105.707		
V	51	105.131		
Cr	52	104.998		
Cr	53			
Mn	55	106.965		
Co	59	108.884		
Ni	60	103.148		
Cu	65	102.996		
Zn	66	102.971		
Ge	72		108.702	
As	75	101.656		
Se	82	103.863		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:44:45
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	97.671		
Ag	107	97.570		
Cd	111	98.608		
Cd	114			
In	115			111.641
Sn	118	97.807		
Sb	123	95.721		
Ba	135	94.691		
Ce	140			
Tb	159			
Ho	165			
Tl	203	100.265		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.810		
U	238	100.773		
Bi	209			107.433
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:44:45
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 15:45:40
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76618.3	2.3				ug/L	74699	Standard
	Be	9	25.0	40.0	**0.0352**	0.010	29.8	ug/L	8	Standard
	Al	27	2451.9	23.9	**0.0029**	0.004	131.5	ug/L	1493	Standard
	Sc	45	64234.2	1.5				ug/L	61029	Standard
	Ti	47	38.0	9.5	**-0.0165**	0.010	60.4	ug/L	35	Standard
	V	51	1363.5	4.5	**-0.0292**	0.006	19.9	ug/L	1435	Standard
	Cr	52	7102.3	0.4	**-0.1401**	0.016	11.4	ug/L	7635	Standard
	Cr	53	373.3	16.1	**0.0429**	0.063	148.1	ug/L	363	Standard
	Mn	55	1097.0	20.0	**-0.0025**	0.031	1279.7	ug/L	747	Standard
	Co	59	247.3	6.4	**-0.0027**	0.003	101.3	ug/L	234	Standard
	Ni	60	180.3	4.0	**-0.0351**	0.003	7.6	ug/L	215	Standard
	Cu	65	132.0	9.2	**-0.0382**	0.007	18.6	ug/L	104	Standard
	Zn	66	415.3	4.4	**-0.1038**	0.019	18.2	ug/L	363	Standard
>	**Ge**	72	537099.7	1.7				ug/L	495142	Standard
	As	75	-44.5	23.9	**0.0230**	0.008	35.8	ug/L	-50	Standard
	Se	82	14.3	4.3	**-0.0180**	0.004	23.6	ug/L	14	Standard
	Se-1	77	59.7	7.9	**0.1303**	0.047	35.8	ug/L	42	Standard
>	**Ga**	71	51.7	29.6				mg/L	42	Standard
	Rb	85	35.0	0.0				ug/L	17	Standard
	Y	89	479758.9	0.4				ug/L	417098	Standard
>	**Rh**	103	8.3	34.6				ug/L	13	Standard
	Mo	98	277.2	29.2	**0.0885**	0.026	29.3	ug/L	11	Standard
	Ag	107	131.7	17.8	**0.0055**	0.003	58.6	ug/L	68	Standard
	Cd	111	17.2	29.5	**-0.0005**	0.002	465.0	mg/L	2	Standard
	Cd	114	40.7	69.7	**0.0051**	0.005	98.6	ug/L	12	Standard
>	**In**	115	646048.6	1.6				ug/L	587441	Standard
	Sn	118	843.4	4.5	**0.0366**	0.007	19.0	ug/L	497	Standard
	Sb	123	501.9	29.2	**0.0939**	0.030	32.1	ug/L	78	Standard
	Ba	135	52.0	15.0	**0.0071**	0.003	45.6	ug/L	24	Standard
	Ce	140	70.0	58.5				ug/L	25	Standard
>	**Tb**	159	979678.9	1.9				ug/L	939701	Standard
	Ho	165	13.3	114.6				ug/L	10	Standard
	Tl	203	68.0	38.6	**0.0062**	0.003	47.0	ug/L	4	Standard
	Tl	205	56.7	27.0	**0.0078**	0.002	25.4	ug/L	2	Standard
	Pb	206	311.7	5.0	**-0.0149**	0.002	16.8	ug/L	240	Standard
	Pb	207	257.0	6.3	**-0.0177**	0.002	12.3	ug/L	198	Standard
	Pb	208	1140.4	10.1	**-0.0154**	0.005	31.9	ug/L	883	Standard
	U	238	55.7	48.1	**0.0033**	0.004	107.8	ug/L	3	Standard
>	**Bi**	209	542619.7	2.4				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:47:57
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.9504**	31.082	173.2	mg/L	0	Standard
Mg	24	40.0	54.5	**-0.0162**	0.026	161.8	mg/L	37	Standard
K	39	53.3	59.5	**-0.0020**	0.067	3445.3	mg/L	38	Standard
Ca	43	18.3	103.3	**-1.4157**	1.156	81.6	mg/L	10	Standard
Fe	54	130.7	11.8	**-0.0329**	0.014	41.5	mg/L	179	Standard
Fe	57	301.7	9.1	**-0.1005**	0.081	80.3	mg/L	335	Standard
> **Sc-1**	45	64234.2	1.5				mg/L	61029	Standard
Cl	35	26295.5	2.8				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1633.4	7.7				ug/L	1547	Standard
P	31	19669.4	2.1				ug/L	15286	Standard
S	34	2230.2	5.1				ug/L	1572	Standard
Sr	88	70.0	21.4				ug/L	73	Standard
C	12	46.7	61.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	12.2	49.4				mg/L	6	Standard
Ho-1	165	13.3	114.6				mg/L	10	Standard
Er	166	23.3	107.9				mg/L	13	Standard
I	127	1980.1	22.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.474	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:47:57
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.977
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.556
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:47:57
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBW 60 WG574001-04
Sample Date/Time: Wednesday, June 29, 2016 15:48:53
Number of Replicates: 3
Autosampler Position: 205
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79282.0	2.9				ug/L	74699	Standard
	Be	9	31.7	134.3	**0.0407**	0.044	108.3	ug/L	8	Standard
	Al	27	19345.2	59.5	**0.1093**	0.069	63.4	ug/L	1493	Standard
	Sc	45	67500.3	4.8				ug/L	61029	Standard
	Ti	47	51.7	27.5	**0.0181**	0.032	177.8	ug/L	35	Standard
	V	51	1877.4	33.1	**0.0362**	0.072	198.6	ug/L	1435	Standard
	Cr	52	9295.2	4.7	**0.1500**	0.032	21.3	ug/L	7635	Standard
	Cr	53	908.4	6.4	**0.5905**	0.063	10.7	ug/L	363	Standard
	Mn	55	10971.3	111.8	**1.3060**	1.603	122.8	ug/L	747	Standard
	Co	59	555.3	78.4	**0.0407**	0.060	146.5	ug/L	234	Standard
	Ni	60	485.3	47.2	**0.1173**	0.109	92.6	ug/L	215	Standard
	Cu	65	307.3	25.1	**0.0504**	0.035	70.1	ug/L	104	Standard
	Zn	66	1901.5	5.5	**1.1744**	0.056	4.8	ug/L	363	Standard
>	**Ge**	72	528898.5	2.0				ug/L	495142	Standard
	As	75	-0.1	63053.1	**0.0590**	0.043	73.0	ug/L	-50	Standard
	Se	82	20.7	41.6	**0.0373**	0.070	187.2	ug/L	14	Standard
	Se-1	77	69.7	5.4	**0.2713**	0.061	22.4	ug/L	42	Standard
>	**Ga**	71	70.0	75.3				mg/L	42	Standard
	Rb	85	391.7	126.9				ug/L	17	Standard
	Y	89	488721.1	2.3				ug/L	417098	Standard
>	**Rh**	103	16.7	105.4				ug/L	13	Standard
	Mo	98	140.7	71.4	**0.0402**	0.032	80.6	ug/L	11	Standard
	Ag	107	241.3	88.8	**0.0193**	0.028	147.0	ug/L	68	Standard
	Cd	111	97.4	139.8	**0.0324**	0.056	173.4	mg/L	2	Standard
	Cd	114	250.4	134.1	**0.0402**	0.057	140.7	ug/L	12	Standard
>	**In**	115	673056.2	1.0				ug/L	587441	Standard
	Sn	118	918.4	10.7	**0.0424**	0.014	32.3	ug/L	497	Standard
	Sb	123	377.5	33.3	**0.0663**	0.023	34.5	ug/L	78	Standard
	Ba	135	710.7	130.6	**0.2849**	0.392	137.5	ug/L	24	Standard
	Ce	140	2190.4	144.6				ug/L	25	Standard
>	**Tb**	159	1033072.6	2.9				ug/L	939701	Standard
	Ho	165	46.7	118.0				ug/L	10	Standard
	Tl	203	354.0	129.7	**0.0375**	0.050	134.0	ug/L	4	Standard
	Tl	205	175.0	119.3	**0.0226**	0.026	116.3	ug/L	2	Standard
	Pb	206	584.7	65.9	**0.0302**	0.064	211.0	ug/L	240	Standard
	Pb	207	508.0	67.6	**0.0291**	0.064	219.9	ug/L	198	Standard
	Pb	208	1976.7	53.7	**0.0238**	0.049	207.9	ug/L	883	Standard
	U	238	147.7	117.1	**0.0152**	0.023	148.4	ug/L	3	Standard
>	**Bi**	209	558931.4	2.9				ug/L	523984	Standard

Sample ID: PBW 60 WG574001-04
Report Date/Time: Wednesday, June 29, 2016 15:51:10
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	38.3	32.8	**-0.0206**	0.014	67.5	mg/L	37	Standard
K	39	46.7	37.6	**-0.0214**	0.032	149.6	mg/L	38	Standard
Ca	43	20.0	0.0	**-1.3794**	0.055	4.0	mg/L	10	Standard
Fe	54	191.0	21.0	**0.0073**	0.024	325.9	mg/L	179	Standard
Fe	57	261.7	19.2	**-0.2718**	0.132	48.6	mg/L	335	Standard
> **Sc-1**	45	67500.3	4.8				mg/L	61029	Standard
Cl	35	25962.2	2.0				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	2216.8	6.9				ug/L	1547	Standard
P	31	19524.2	2.3				ug/L	15286	Standard
S	34	2248.5	4.1				ug/L	1572	Standard
Sr	88	73.3	27.6				ug/L	73	Standard
C	12	46.7	44.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	67.1	124.3				mg/L	6	Standard
Ho-1	165	46.7	118.0				mg/L	10	Standard
Er	166	60.0	130.2				mg/L	13	Standard
I	127	1978.5	35.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		106.136	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		106.817	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBW 60 WG574001-04
Report Date/Time: Wednesday, June 29, 2016 15:51:10
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.574
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.669
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 60 WG574001-04
Report Date/Time: Wednesday, June 29, 2016 15:51:10
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSW 60 WG574001-05
Sample Date/Time: Wednesday, June 29, 2016 15:52:04
Number of Replicates: 3
Autosampler Position: 206
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	78008.7	2.7				ug/L	74699	Standard
	Be	9	49798.4	3.1	**54.1693**	0.583	1.1	ug/L	8	Standard
	Al	27	8265.6	3.2	**0.0405**	0.001	2.7	ug/L	1493	Standard
	Sc	45	64878.8	4.3				ug/L	61029	Standard
	Ti	47	43.0	16.3	**-0.0042**	0.019	444.7	ug/L	35	Standard
	V	51	424123.5	0.9	**51.5862**	0.653	1.3	ug/L	1435	Standard
	Cr	52	430096.1	0.4	**52.4970**	1.185	2.3	ug/L	7635	Standard
	Cr	53	53307.0	1.3	**52.8599**	1.230	2.3	ug/L	363	Standard
	Mn	55	405428.0	0.3	**53.5518**	1.000	1.9	ug/L	747	Standard
	Co	59	387903.4	0.1	**54.1349**	1.154	2.1	ug/L	234	Standard
	Ni	60	106395.6	1.4	**52.2826**	1.359	2.6	ug/L	215	Standard
	Cu	65	107418.5	1.1	**53.1215**	1.283	2.4	ug/L	104	Standard
	Zn	66	63602.2	0.4	**53.2433**	1.221	2.3	ug/L	363	Standard
>	Ge	72	537059.3	2.0				ug/L	495142	Standard
	As	75	62286.8	1.3	**51.3419**	1.295	2.5	ug/L	-50	Standard
	Se	82	6132.0	1.6	**51.3807**	0.931	1.8	ug/L	14	Standard
	Se-1	77	4215.6	1.2	**52.6015**	0.951	1.8	ug/L	42	Standard
>	Ga	71	76.7	32.8				mg/L	42	Standard
	Rb	85	50.0	36.1				ug/L	17	Standard
	Y	89	473307.8	1.1				ug/L	417098	Standard
>	Rh	103	23.3	75.3				ug/L	13	Standard
	Mo	98	59.9	20.8	**0.0148**	0.004	26.5	ug/L	11	Standard
	Ag	107	358191.0	0.9	**50.3679**	1.706	3.4	ug/L	68	Standard
	Cd	111	112882.0	2.3	**49.4684**	1.332	2.7	mg/L	2	Standard
	Cd	114	273909.8	1.0	**48.9526**	1.484	3.0	ug/L	12	Standard
>	In	115	641215.1	2.7				ug/L	587441	Standard
	Sn	118	785.0	7.1	**0.0286**	0.011	37.9	ug/L	497	Standard
	Sb	123	249794.4	1.0	**49.0383**	1.564	3.2	ug/L	78	Standard
	Ba	135	107708.9	1.4	**48.2176**	1.451	3.0	ug/L	24	Standard
	Ce	140	101.7	27.1				ug/L	25	Standard
>	Tb	159	990038.9	2.7				ug/L	939701	Standard
	Ho	165	11.7	99.0				ug/L	10	Standard
	Tl	203	439976.8	0.4	**51.4932**	0.969	1.9	ug/L	4	Standard
	Tl	205	398605.3	0.6	**53.7491**	0.791	1.5	ug/L	2	Standard
	Pb	206	290681.6	0.4	**52.4241**	0.636	1.2	ug/L	240	Standard
	Pb	207	247225.7	0.4	**50.1841**	0.852	1.7	ug/L	198	Standard
	Pb	208	1030904.4	0.6	**52.9701**	1.178	2.2	ug/L	883	Standard
	U	238	368133.3	0.5	**51.5924**	0.851	1.6	ug/L	3	Standard
>	Bi	209	543222.2	1.6				ug/L	523984	Standard

Sample ID: LCSW 60 WG574001-05
Report Date/Time: Wednesday, June 29, 2016 15:54:21
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	35.0	28.6	**-0.0227**	0.012	52.0	mg/L	37	Standard
K	39	28.3	50.9	**-0.0559**	0.027	48.5	mg/L	38	Standard
Ca	43	35.0	28.6	**-0.4579**	0.506	110.5	mg/L	10	Standard
Fe	54	138.7	22.6	**-0.0277**	0.025	88.8	mg/L	179	Standard
Fe	57	243.3	16.6	**-0.2916**	0.161	55.3	mg/L	335	Standard
> **Sc-1**	45	64878.8	4.3				mg/L	61029	Standard
Cl	35	24949.9	2.3				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	1880.1	6.1				ug/L	1547	Standard
P	31	18362.8	1.6				ug/L	15286	Standard
S	34	2046.8	4.1				ug/L	1572	Standard
Sr	88	70.0	18.9				ug/L	73	Standard
C	12	56.7	56.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	2.2	249.0				mg/L	6	Standard
Ho-1	165	11.7	99.0				mg/L	10	Standard
Er	166	23.3	49.5				mg/L	13	Standard
I	127	1781.8	25.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		104.431	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.466	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 60 WG574001-05
Report Date/Time: Wednesday, June 29, 2016 15:54:21
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.154
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.671
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 60 WG574001-05
Report Date/Time: Wednesday, June 29, 2016 15:54:21
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124702 WG574001-01
Sample Date/Time: Wednesday, June 29, 2016 15:55:16
Number of Replicates: 3
Autosampler Position: 207
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	134635.3	6.0				ug/L	74699	Standard
	Be	9	80.0	18.8	**0.0585**	0.013	21.4	ug/L	8	Standard
	Al	27	38117257.0	1.5	**144.1796**	7.359	5.1	ug/L	1493	Standard
	Sc	45	57575.9	1.8				ug/L	61029	Standard
	Ti	47	164.0	8.6	**0.3117**	0.039	12.5	ug/L	35	Standard
	V	51	2694.7	7.4	**0.1512**	0.027	17.7	ug/L	1435	Standard
	Cr	52	13340.2	1.7	**0.7256**	0.017	2.4	ug/L	7635	Standard
	Cr	53	1556.7	8.2	**1.3078**	0.145	11.1	ug/L	363	Standard
	Mn	55	855520.0	4.5	**119.2592**	4.647	3.9	ug/L	747	Standard
	Co	59	4433.3	4.9	**0.6151**	0.027	4.4	ug/L	234	Standard
	Ni	60	11360.3	1.4	**5.7737**	0.123	2.1	ug/L	215	Standard
	Cu	65	4582.0	1.3	**2.2889**	0.015	0.7	ug/L	104	Standard
	Zn	66	3866.8	1.4	**2.9861**	0.052	1.7	ug/L	363	Standard
>	**Ge**	72	509457.0	0.7				ug/L	495142	Standard
	As	75	332.4	10.8	**0.3479**	0.030	8.6	ug/L	-50	Standard
	Se	82	161.9	7.9	**1.2952**	0.108	8.4	ug/L	14	Standard
	Se-1	77	102.3	14.4	**0.7391**	0.192	26.0	ug/L	42	Standard
>	**Ga**	71	95.0	18.2				mg/L	42	Standard
	Rb	85	141592.2	2.6				ug/L	17	Standard
	Y	89	417616.8	3.6				ug/L	417098	Standard
>	**Rh**	103	198.3	13.9				ug/L	13	Standard
	Mo	98	173.0	3.6	**0.0623**	0.003	5.6	ug/L	11	Standard
	Ag	107	339.3	7.8	**0.0417**	0.005	13.0	ug/L	68	Standard
	Cd	111	139.7	8.2	**0.0622**	0.007	10.5	mg/L	2	Standard
	Cd	114	358.4	9.8	**0.0712**	0.009	12.2	ug/L	12	Standard
>	**In**	115	560534.9	2.1				ug/L	587441	Standard
	Sn	118	620.0	5.6	**0.0166**	0.008	49.6	ug/L	497	Standard
	Sb	123	641.7	15.5	**0.1401**	0.024	17.2	ug/L	78	Standard
	Ba	135	5098.9	2.7	**2.5958**	0.100	3.9	ug/L	24	Standard
	Ce	140	815.0	33.2				ug/L	25	Standard
>	**Tb**	159	928792.3	1.3				ug/L	939701	Standard
	Ho	165	11.7	65.5				ug/L	10	Standard
	Tl	203	1143.7	8.4	**0.1440**	0.014	9.4	ug/L	4	Standard
	Tl	205	1088.4	7.4	**0.1600**	0.012	7.4	ug/L	2	Standard
	Pb	206	578.7	9.2	**0.0425**	0.010	23.1	ug/L	240	Standard
	Pb	207	493.0	9.5	**0.0393**	0.012	31.8	ug/L	198	Standard
	Pb	208	1938.4	3.0	**0.0345**	0.004	12.8	ug/L	883	Standard
	U	238	1219.0	4.3	**0.1816**	0.009	5.1	ug/L	3	Standard
>	**Bi**	209	499083.1	3.3				ug/L	523984	Standard

Sample ID: L1606124702 WG574001-01
Report Date/Time: Wednesday, June 29, 2016 15:57:33
Page 1

Approved: July 04, 2016

Na	23	20.0	43.3	**245.3647**	110.079	44.9	mg/L	0	Standard
Mg	24	80086.1	0.7	**109.5385**	2.239	2.0	mg/L	37	Standard
K	39	1376.7	6.5	**3.1098**	0.163	5.2	mg/L	38	Standard
Ca	43	2828.6	6.5	**186.7611**	8.963	4.8	mg/L	10	Standard
Fe	54	188.7	25.7	**0.0318**	0.047	146.5	mg/L	179	Standard
Fe	57	1751.8	4.9	**5.2348**	0.203	3.9	mg/L	335	Standard
> **Sc-1**	45	57575.9	1.8				mg/L	61029	Standard
Cl	35	65110.3	1.2				ug/L	20195	Standard
Kr	83	3.7	68.6				ug/L	4	Standard
Br	81	22346.5	6.5				ug/L	1547	Standard
P	31	17116.3	3.4				ug/L	15286	Standard
S	34	2223.5	6.9				ug/L	1572	Standard
Sr	88	836.7	13.9				ug/L	73	Standard
C	12	76.7	74.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	52.1	77.4				mg/L	6	Standard
Ho-1	165	11.7	65.5				mg/L	10	Standard
Er	166	26.7	57.3				mg/L	13	Standard
I	127	41219.6	7.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		180.238	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		102.891	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606124702 WG574001-01
Report Date/Time: Wednesday, June 29, 2016 15:57:33
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.420
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.248
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124702DP WG574001-06
Sample Date/Time: Wednesday, June 29, 2016 15:58:27
Number of Replicates: 3
Autosampler Position: 208
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	151702.5	2.0				ug/L	74699	Standard
	Be	9	51.7	59.1	**0.0368**	0.018	48.5	ug/L	8	Standard
	Al	27	41839885.7	0.1	**140.2128**	2.854	2.0	ug/L	1493	Standard
	Sc	45	56811.4	3.1				ug/L	61029	Standard
	Ti	47	131.7	6.1	**0.2545**	0.026	10.2	ug/L	35	Standard
	V	51	2376.3	9.9	**0.1332**	0.026	19.9	ug/L	1435	Standard
	Cr	52	12951.9	1.5	**0.8048**	0.052	6.4	ug/L	7635	Standard
	Cr	53	1565.1	1.7	**1.4417**	0.042	2.9	ug/L	363	Standard
	Mn	55	867683.1	1.0	**130.2493**	3.502	2.7	ug/L	747	Standard
	Co	59	4344.6	3.7	**0.6509**	0.022	3.4	ug/L	234	Standard
	Ni	60	12342.7	0.7	**6.7729**	0.126	1.9	ug/L	215	Standard
	Cu	65	4552.0	1.4	**2.4548**	0.036	1.5	ug/L	104	Standard
	Zn	66	4125.6	2.3	**3.4960**	0.055	1.6	ug/L	363	Standard
>	**Ge**	72	473381.7	2.0				ug/L	495142	Standard
	As	75	389.2	1.9	**0.4231**	0.002	0.6	ug/L	-50	Standard
	Se	82	181.1	4.2	**1.5888**	0.099	6.2	ug/L	14	Standard
	Se-1	77	106.3	18.0	**0.8981**	0.248	27.6	ug/L	42	Standard
>	**Ga**	71	78.3	22.4				mg/L	42	Standard
	Rb	85	147149.5	0.7				ug/L	17	Standard
	Y	89	405581.1	1.8				ug/L	417098	Standard
>	**Rh**	103	178.3	29.5				ug/L	13	Standard
	Mo	98	220.6	12.3	**0.0817**	0.013	15.7	ug/L	11	Standard
	Ag	107	177.0	68.0	**0.0160**	0.020	127.1	ug/L	68	Standard
	Cd	111	65.9	51.9	**0.0256**	0.018	71.0	mg/L	2	Standard
	Cd	114	151.1	49.7	**0.0291**	0.016	55.4	ug/L	12	Standard
>	**In**	115	557347.0	2.4				ug/L	587441	Standard
	Sn	118	766.7	0.4	**0.0437**	0.004	8.6	ug/L	497	Standard
	Sb	123	305.8	17.0	**0.0650**	0.013	20.6	ug/L	78	Standard
	Ba	135	5047.2	3.4	**2.5853**	0.149	5.7	ug/L	24	Standard
	Ce	140	491.7	100.3				ug/L	25	Standard
>	**Tb**	159	935324.5	2.7				ug/L	939701	Standard
	Ho	165	18.3	103.3				ug/L	10	Standard
	Tl	203	947.0	9.8	**0.1179**	0.014	11.5	ug/L	4	Standard
	Tl	205	846.7	27.0	**0.1237**	0.035	28.6	ug/L	2	Standard
	Pb	206	418.3	37.6	**0.0106**	0.032	302.9	ug/L	240	Standard
	Pb	207	351.7	32.0	**0.0074**	0.026	354.3	ug/L	198	Standard
	Pb	208	1404.0	30.0	**0.0041**	0.025	608.6	ug/L	883	Standard
	U	238	972.4	9.3	**0.1427**	0.016	11.1	ug/L	3	Standard
>	**Bi**	209	503603.8	1.5				ug/L	523984	Standard

Sample ID: L1606124702DP WG574001-06
Report Date/Time: Wednesday, June 29, 2016 16:00:44
Page 1

Approved: July 04, 2016

Na	23	25.0	34.6	**311.2656**	112.939	36.3	mg/L	0	Standard
Mg	24	85343.0	0.9	**118.3682**	4.384	3.7	mg/L	37	Standard
K	39	1298.4	2.9	**2.9712**	0.147	4.9	mg/L	38	Standard
Ca	43	2830.3	7.5	**189.4839**	12.043	6.4	mg/L	10	Standard
Fe	54	146.6	14.0	**-0.0044**	0.021	483.2	mg/L	179	Standard
Fe	57	1698.4	7.8	**5.1376**	0.623	12.1	mg/L	335	Standard
> **Sc-1**	45	56811.4	3.1				mg/L	61029	Standard
Cl	35	65999.5	1.4				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	22520.1	6.6				ug/L	1547	Standard
P	31	16799.3	3.7				ug/L	15286	Standard
S	34	2040.1	3.0				ug/L	1572	Standard
Sr	88	920.0	8.9				ug/L	73	Standard
C	12	83.3	25.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	173.2				mg/L	3	Standard
Dy	164	32.1	98.6				mg/L	6	Standard
Ho-1	165	18.3	103.3				mg/L	10	Standard
Er	166	26.7	43.3				mg/L	13	Standard
I	127	45859.6	7.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		203.086	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.605	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606124702DP WG574001-06
Report Date/Time: Wednesday, June 29, 2016 16:00:44
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.877
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.110
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124702DP WG574001-06
Report Date/Time: Wednesday, June 29, 2016 16:00:44
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126501 WG574001-02
Sample Date/Time: Wednesday, June 29, 2016 16:01:39
Number of Replicates: 3
Autosampler Position: 209
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	112017.9	3.5				ug/L	74699	Standard
	Be	9	21.7	35.3	**0.0242**	0.006	26.3	ug/L	8	Standard
	Al	27	58851331.1	1.0	**267.2037**	7.701	2.9	ug/L	1493	Standard
	Sc	45	63265.2	2.1				ug/L	61029	Standard
	Ti	47	314.3	6.0	**0.7250**	0.050	6.9	ug/L	35	Standard
	V	51	84379.6	0.6	**11.0401**	0.098	0.9	ug/L	1435	Standard
	Cr	52	25344.2	1.7	**2.4155**	0.065	2.7	ug/L	7635	Standard
	Cr	53	30644.3	3.7	**33.0161**	1.295	3.9	ug/L	363	Standard
	Mn	55	3058.3	0.3	**0.2937**	0.003	0.9	ug/L	747	Standard
	Co	59	981.4	1.2	**0.1122**	0.002	1.4	ug/L	234	Standard
	Ni	60	4598.7	1.8	**2.3463**	0.051	2.2	ug/L	215	Standard
	Cu	65	1226.7	2.1	**0.5592**	0.015	2.8	ug/L	104	Standard
	Zn	66	2978.0	3.2	**2.2873**	0.091	4.0	ug/L	363	Standard
>	**Ge**	72	492356.2	0.3				ug/L	495142	Standard
	As	75	3467.1	5.3	**3.1728**	0.175	5.5	ug/L	-50	Standard
	Se	82	1561.8	3.2	**14.1732**	0.483	3.4	ug/L	14	Standard
	Se-1	77	3085.3	7.4	**41.8633**	3.228	7.7	ug/L	42	Standard
>	**Ga**	71	63.3	27.7				mg/L	42	Standard
	Rb	85	9206.1	3.0				ug/L	17	Standard
	Y	89	442169.1	3.5				ug/L	417098	Standard
>	**Rh**	103	255.0	11.9				ug/L	13	Standard
	Mo	98	5205.2	3.3	**1.8993**	0.033	1.7	ug/L	11	Standard
	Ag	107	803.7	5.3	**0.1074**	0.005	4.8	ug/L	68	Standard
	Cd	111	45.0	9.2	**0.0131**	0.002	16.6	mg/L	2	Standard
	Cd	114	163.3	22.8	**0.0289**	0.007	22.9	ug/L	12	Standard
>	**In**	115	602229.8	3.0				ug/L	587441	Standard
	Sn	118	683.3	9.8	**0.0193**	0.009	45.4	ug/L	497	Standard
	Sb	123	210.7	19.8	**0.0397**	0.008	19.9	ug/L	78	Standard
	Ba	135	21407.4	0.7	**10.1914**	0.253	2.5	ug/L	24	Standard
	Ce	140	328.3	10.1				ug/L	25	Standard
>	**Tb**	159	984916.0	0.2				ug/L	939701	Standard
	Ho	165	35.0	24.7				ug/L	10	Standard
	Tl	203	94.3	9.0	**0.0098**	0.001	12.0	ug/L	4	Standard
	Tl	205	65.0	40.7	**0.0094**	0.004	40.6	ug/L	2	Standard
	Pb	206	401.3	3.0	**0.0049**	0.001	28.2	ug/L	240	Standard
	Pb	207	354.0	5.2	**0.0056**	0.003	55.2	ug/L	198	Standard
	Pb	208	1436.0	3.5	**0.0035**	0.003	74.9	ug/L	883	Standard
	U	238	31993.4	1.5	**4.7067**	0.092	1.9	ug/L	3	Standard
>	**Bi**	209	517006.9	1.2				ug/L	523984	Standard

Sample ID: L1606126501 WG574001-02
Report Date/Time: Wednesday, June 29, 2016 16:03:56
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**55.9813**	56.756	101.4	mg/L	0	Standard
Mg	24	180057.1	1.6	**224.1644**	1.858	0.8	mg/L	37	Standard
K	39	641.7	5.8	**1.2549**	0.096	7.6	mg/L	38	Standard
Ca	43	708.3	7.0	**40.6474**	3.060	7.5	mg/L	10	Standard
Fe	54	157.9	5.8	**-0.0091**	0.008	86.4	mg/L	179	Standard
Fe	57	550.0	17.3	**0.7314**	0.333	45.5	mg/L	335	Standard
> **Sc-1**	45	63265.2	2.1				mg/L	61029	Standard
Cl	35	33676.1	1.7				ug/L	20195	Standard
Kr	83	0.3	173.2				ug/L	4	Standard
Br	81	48808.6	5.5				ug/L	1547	Standard
P	31	23084.3	3.5				ug/L	15286	Standard
S	34	2106.8	4.3				ug/L	1572	Standard
Sr	88	1195.0	6.7				ug/L	73	Standard
C	12	130.0	40.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	9.0	113.3				mg/L	6	Standard
Ho-1	165	35.0	24.7				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	49504.2	3.9				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		149.959	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		99.437	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606126501 WG574001-02
Report Date/Time: Wednesday, June 29, 2016 16:03:56
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	102.517	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	98.668	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126502S WG574001-07
Sample Date/Time: Wednesday, June 29, 2016 16:04:50
Number of Replicates: 3
Autosampler Position: 210
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	109865.8	4.5				ug/L	74699	Standard
	Be	9	61395.8	1.0	**47.4912**	2.348	4.9	ug/L	8	Standard
	Al	27	61297739.1	1.7	**284.0961**	17.361	6.1	ug/L	1493	Standard
	Sc	45	64222.6	2.9				ug/L	61029	Standard
	Ti	47	334.7	6.1	**0.7462**	0.065	8.7	ug/L	35	Standard
	V	51	514269.3	0.9	**65.7128**	0.488	0.7	ug/L	1435	Standard
	Cr	52	414721.1	1.2	**53.1400**	0.316	0.6	ug/L	7635	Standard
	Cr	53	79122.8	2.2	**82.5246**	0.802	1.0	ug/L	363	Standard
	Mn	55	379075.2	0.4	**52.5566**	0.801	1.5	ug/L	747	Standard
	Co	59	375122.1	1.7	**54.9533**	1.379	2.5	ug/L	234	Standard
	Ni	60	98886.4	1.8	**50.9941**	0.435	0.9	ug/L	215	Standard
	Cu	65	94781.9	0.9	**49.1885**	0.428	0.9	ug/L	104	Standard
	Zn	66	59708.1	1.0	**52.4531**	0.349	0.7	ug/L	363	Standard
>	Ge	72	511580.8	1.5				ug/L	495142	Standard
	As	75	64226.1	0.6	**55.5617**	0.582	1.0	ug/L	-50	Standard
	Se	82	7637.7	1.0	**67.2251**	1.117	1.7	ug/L	14	Standard
	Se-1	77	7177.4	2.9	**94.4764**	1.329	1.4	ug/L	42	Standard
>	Ga	71	96.7	7.9				mg/L	42	Standard
	Rb	85	9571.4	0.6				ug/L	17	Standard
	Y	89	459795.2	1.7				ug/L	417098	Standard
>	Rh	103	293.3	12.0				ug/L	13	Standard
	Mo	98	5421.1	3.7	**1.9704**	0.066	3.4	ug/L	11	Standard
	Ag	107	318338.0	1.0	**47.4485**	0.356	0.8	ug/L	68	Standard
	Cd	111	104991.2	1.0	**48.7852**	0.464	1.0	mg/L	2	Standard
	Cd	114	250399.7	1.5	**47.4405**	0.760	1.6	ug/L	12	Standard
>	In	115	604563.1	0.3				ug/L	587441	Standard
	Sn	118	798.4	4.4	**0.0381**	0.006	16.0	ug/L	497	Standard
	Sb	123	240747.3	1.0	**50.1002**	0.397	0.8	ug/L	78	Standard
	Ba	135	122571.5	0.3	**58.1746**	0.379	0.7	ug/L	24	Standard
	Ce	140	328.3	0.9				ug/L	25	Standard
>	Tb	159	981221.4	1.4				ug/L	939701	Standard
	Ho	165	28.3	44.4				ug/L	10	Standard
	Tl	203	414235.7	0.7	**51.0974**	0.645	1.3	ug/L	4	Standard
	Tl	205	378219.2	0.6	**53.7534**	0.090	0.2	ug/L	2	Standard
	Pb	206	277784.6	0.1	**52.8053**	0.305	0.6	ug/L	240	Standard
	Pb	207	237018.9	0.5	**50.7105**	0.514	1.0	ug/L	198	Standard
	Pb	208	965977.0	0.4	**52.3096**	0.506	1.0	ug/L	883	Standard
	U	238	407586.9	0.7	**60.2072**	0.700	1.2	ug/L	3	Standard
>	Bi	209	515321.0	0.6				ug/L	523984	Standard

Sample ID: L1606126502S WG574001-07
Report Date/Time: Wednesday, June 29, 2016 16:07:07
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.6720**	30.600	173.2	mg/L	0	Standard
Mg	24	189285.1	1.3	**232.3057**	9.548	4.1	mg/L	37	Standard
K	39	641.7	6.1	**1.2328**	0.045	3.6	mg/L	38	Standard
Ca	43	760.0	12.3	**43.0240**	4.391	10.2	mg/L	10	Standard
Fe	54	185.4	16.8	**0.0114**	0.028	245.2	mg/L	179	Standard
Fe	57	630.0	4.2	**0.9612**	0.087	9.0	mg/L	335	Standard
> **Sc-1**	45	64222.6	2.9				mg/L	61029	Standard
Cl	35	34933.6	1.5				ug/L	20195	Standard
Kr	83	2.7	114.6				ug/L	4	Standard
Br	81	53306.9	1.0				ug/L	1547	Standard
P	31	23740.3	1.4				ug/L	15286	Standard
S	34	2118.5	3.3				ug/L	1572	Standard
Sr	88	1145.0	3.3				ug/L	73	Standard
C	12	140.0	21.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	28.6	38.4				mg/L	6	Standard
Ho-1	165	28.3	44.4				mg/L	10	Standard
Er	166	30.0	66.7				mg/L	13	Standard
I	127	81496.0	7.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		147.078	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.320	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126502S WG574001-07
Report Date/Time: Wednesday, June 29, 2016 16:07:07
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	102.915	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	98.347	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126503SD WG574001-08
Sample Date/Time: Wednesday, June 29, 2016 16:08:02
Number of Replicates: 3
Autosampler Position: 211
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	108017.3	3.5				ug/L	74699	Standard
	Be	9	61104.9	5.2	**47.9870**	0.846	1.8	ug/L	8	Standard
	Al	27	60436124.3	2.1	**284.5445**	9.073	3.2	ug/L	1493	Standard
	Sc	45	65195.1	3.6				ug/L	61029	Standard
	Ti	47	321.3	5.9	**0.6738**	0.054	8.0	ug/L	35	Standard
	V	51	521782.6	1.8	**63.5667**	1.206	1.9	ug/L	1435	Standard
	Cr	52	431117.5	0.3	**52.6702**	0.983	1.9	ug/L	7635	Standard
	Cr	53	81998.0	1.1	**81.5509**	0.729	0.9	ug/L	363	Standard
	Mn	55	379421.8	1.7	**50.1647**	1.838	3.7	ug/L	747	Standard
	Co	59	378746.6	2.0	**52.9167**	2.098	4.0	ug/L	234	Standard
	Ni	60	101643.5	0.7	**49.9872**	1.144	2.3	ug/L	215	Standard
	Cu	65	96036.4	1.2	**47.5169**	0.455	1.0	ug/L	104	Standard
	Zn	66	60629.4	1.5	**50.7689**	0.479	0.9	ug/L	363	Standard
>	Ge	72	536566.5	1.9				ug/L	495142	Standard
	As	75	65668.7	0.7	**54.1701**	0.796	1.5	ug/L	-50	Standard
	Se	82	7769.6	1.1	**65.2136**	1.970	3.0	ug/L	14	Standard
	Se-1	77	7316.4	1.6	**91.8340**	1.824	2.0	ug/L	42	Standard
>	Ga	71	56.7	22.2				mg/L	42	Standard
	Rb	85	9684.8	3.9				ug/L	17	Standard
	Y	89	458590.9	3.1				ug/L	417098	Standard
>	Rh	103	275.0	12.6				ug/L	13	Standard
	Mo	98	5397.9	1.9	**1.9792**	0.040	2.0	ug/L	11	Standard
	Ag	107	315299.7	0.9	**47.4082**	1.008	2.1	ug/L	68	Standard
	Cd	111	105649.0	0.5	**49.5183**	0.751	1.5	mg/L	2	Standard
	Cd	114	256574.5	1.5	**49.0334**	1.028	2.1	ug/L	12	Standard
>	In	115	599465.1	2.0				ug/L	587441	Standard
	Sn	118	820.0	11.4	**0.0428**	0.015	34.4	ug/L	497	Standard
	Sb	123	245153.5	0.2	**51.4644**	0.950	1.8	ug/L	78	Standard
	Ba	135	122768.4	0.8	**58.7829**	1.524	2.6	ug/L	24	Standard
	Ce	140	333.3	10.5				ug/L	25	Standard
>	Tb	159	983758.7	0.8				ug/L	939701	Standard
	Ho	165	18.3	41.7				ug/L	10	Standard
	Tl	203	416470.0	0.7	**50.8907**	0.470	0.9	ug/L	4	Standard
	Tl	205	377588.7	0.1	**53.1619**	0.172	0.3	ug/L	2	Standard
	Pb	206	279930.3	0.5	**52.7144**	0.351	0.7	ug/L	240	Standard
	Pb	207	238634.3	0.7	**50.5762**	0.275	0.5	ug/L	198	Standard
	Pb	208	958199.8	1.0	**51.4011**	0.688	1.3	ug/L	883	Standard
	U	238	405052.0	0.8	**59.2716**	0.671	1.1	ug/L	3	Standard
>	Bi	209	520189.0	0.3				ug/L	523984	Standard

Sample ID: L1606126503SD WG574001-08
Report Date/Time: Wednesday, June 29, 2016 16:10:18
Page 1

Approved: July 04, 2016

Na	23	5.0	173.2	**51.9909**	90.042	173.2	mg/L	0	Standard
Mg	24	184795.4	2.1	**223.5320**	12.550	5.6	mg/L	37	Standard
K	39	668.3	13.7	**1.2668**	0.164	13.0	mg/L	38	Standard
Ca	43	725.0	7.3	**40.3567**	2.975	7.4	mg/L	10	Standard
Fe	54	151.4	21.3	**-0.0183**	0.024	132.3	mg/L	179	Standard
Fe	57	623.3	16.5	**0.9045**	0.268	29.6	mg/L	335	Standard
> **Sc-1**	45	65195.1	3.6				mg/L	61029	Standard
Cl	35	33686.7	1.5				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	51370.3	3.2				ug/L	1547	Standard
P	31	23124.3	1.4				ug/L	15286	Standard
S	34	2101.8	2.3				ug/L	1572	Standard
Sr	88	1251.7	9.6				ug/L	73	Standard
C	12	163.3	19.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	9.4	104.3				mg/L	6	Standard
Ho-1	165	18.3	41.7				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	90519.1	8.2				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		144.604	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		108.366	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606126503SD WG574001-08
Report Date/Time: Wednesday, June 29, 2016 16:10:18
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	102.047
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	99.276
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606126503SD WG574001-08
Report Date/Time: Wednesday, June 29, 2016 16:10:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701
Sample Date/Time: Wednesday, June 29, 2016 16:11:13
Number of Replicates: 3
Autosampler Position: 212
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	160464.4	3.8				ug/L	74699	Standard
	Be	9	12485.1	1.8	**6.6126**	0.132	2.0	ug/L	8	Standard
	Al	27	129371635.5	1.7	**410.0067**	8.683	2.1	ug/L	1493	Standard
	Sc	45	97873.5	2.0				ug/L	61029	Standard
	Ti	47	12940.2	2.9	**32.4229**	1.439	4.4	ug/L	35	Standard
	V	51	62587.9	0.8	**7.6988**	0.173	2.3	ug/L	1435	Standard
	Cr	52	130472.4	1.2	**15.7449**	0.176	1.1	ug/L	7635	Standard
	Cr	53	17299.9	2.8	**17.4949**	0.102	0.6	ug/L	363	Standard
	Mn	55	53885293.3	1.0	**7372.4056**	142.217	1.9	ug/L	747	Standard
	Co	59	1486809.2	2.0	**214.4609**	6.754	3.1	ug/L	234	Standard
	Ni	60	953927.1	1.9	**485.1518**	7.673	1.6	ug/L	215	Standard
	Cu	65	14021.8	2.7	**7.0705**	0.048	0.7	ug/L	104	Standard
	Zn	66	1250022.8	1.4	**1089.5042**	11.046	1.0	ug/L	363	Standard
>	**Ge**	72	519921.5	2.2				ug/L	495142	Standard
	As	75	2369.9	1.5	**2.0748**	0.031	1.5	ug/L	-50	Standard
	Se	82	334.2	2.8	**2.7639**	0.140	5.1	ug/L	14	Standard
	Se-1	77	477.7	9.4	**5.6016**	0.492	8.8	ug/L	42	Standard
>	**Ga**	71	3503.7	1.4				mg/L	42	Standard
	Rb	85	155749.9	1.6				ug/L	17	Standard
	Y	89	3051185.4	0.7				ug/L	417098	Standard
>	**Rh**	103	200.0	9.0				ug/L	13	Standard
	Mo	98	648.1	0.8	**0.2277**	0.002	1.0	ug/L	11	Standard
	Ag	107	406.3	10.5	**0.0469**	0.006	13.8	ug/L	68	Standard
	Cd	111	4037.2	2.2	**1.8458**	0.056	3.0	mg/L	2	Standard
	Cd	114	10321.7	4.5	**1.9292**	0.065	3.4	ug/L	12	Standard
>	**In**	115	612067.4	1.8				ug/L	587441	Standard
	Sn	118	936.7	6.3	**0.0593**	0.012	20.5	ug/L	497	Standard
	Sb	123	674.3	5.8	**0.1344**	0.006	4.1	ug/L	78	Standard
	Ba	135	15717.8	1.3	**7.3556**	0.136	1.8	ug/L	24	Standard
	Ce	140	2639736.8	0.7				ug/L	25	Standard
>	**Tb**	159	1213429.7	1.8				ug/L	939701	Standard
	Ho	165	212831.1	0.7				ug/L	10	Standard
	Tl	203	2423.5	3.4	**0.2853**	0.011	3.9	ug/L	4	Standard
	Tl	205	2225.2	6.3	**0.3038**	0.017	5.8	ug/L	2	Standard
	Pb	206	7086.3	0.1	**1.2240**	0.010	0.8	ug/L	240	Standard
	Pb	207	5547.0	2.2	**1.0710**	0.022	2.1	ug/L	198	Standard
	Pb	208	23218.6	1.0	**1.1348**	0.005	0.5	ug/L	883	Standard
	U	238	25740.2	1.8	**3.6463**	0.044	1.2	ug/L	3	Standard
>	**Bi**	209	536711.9	0.7				ug/L	523984	Standard

Sample ID: L1606124701
Report Date/Time: Wednesday, June 29, 2016 16:13:30
Page 1

Approved: July 04, 2016

Na	23	40.0	57.3	**285.3574**	160.270	56.2	mg/L	0	Standard
Mg	24	87773.6	1.9	**70.5914**	1.090	1.5	mg/L	37	Standard
K	39	1211.7	2.8	**1.5562**	0.053	3.4	mg/L	38	Standard
Ca	43	3308.7	4.2	**127.7939**	4.640	3.6	mg/L	10	Standard
Fe	54	180981.2	2.7	**95.5321**	0.892	0.9	mg/L	179	Standard
Fe	57	48873.8	3.4	**102.4969**	2.098	2.0	mg/L	335	Standard
Sc-1	45	97873.5	2.0				mg/L	61029	Standard
Cl	35	64420.7	1.3				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	26873.8	3.7				ug/L	1547	Standard
P	31	52208.1	1.4				ug/L	15286	Standard
S	34	1956.8	2.0				ug/L	1572	Standard
Sr	88	1171.7	10.6				ug/L	73	Standard
C	12	126.7	46.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	329920.1	1.3				mg/L	6	Standard
Ho-1	165	212831.1	0.7				mg/L	10	Standard
Er	166	185158.1	1.8				mg/L	13	Standard
I	127	88143.1	6.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		214.815	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		105.004	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606124701
Report Date/Time: Wednesday, June 29, 2016 16:13:30
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.192
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.429
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701
Report Date/Time: Wednesday, June 29, 2016 16:13:30
Page 3

Approved: July 04, 2016

Co 59 Upper, S, EEE	Co	59
Ni 60 Upper, S, EEE	Ni	60
Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701PS WG574688-01
Sample Date/Time: Wednesday, June 29, 2016 16:14:25
Number of Replicates: 3
Autosampler Position: 213
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	166369.0	3.1				ug/L	74699	Standard
	Be	9	78228.1	1.5	**39.9183**	0.744	1.9	ug/L	8	Standard
	Al	27	127858501.0	1.9	**390.7420**	6.863	1.8	ug/L	1493	Standard
	Sc	45	95075.8	2.0				ug/L	61029	Standard
	Ti	47	12076.5	3.4	**30.6847**	0.840	2.7	ug/L	35	Standard
	V	51	487373.4	0.7	**62.1823**	0.880	1.4	ug/L	1435	Standard
	Cr	52	506749.0	0.5	**65.0704**	0.602	0.9	ug/L	7635	Standard
	Cr	53	63546.3	1.7	**66.1298**	1.201	1.8	ug/L	363	Standard
	Mn	55	53730576.4	0.3	**7459.5309**	74.503	1.0	ug/L	747	Standard
	Co	59	1866113.9	0.2	**273.1251**	3.396	1.2	ug/L	234	Standard
	Ni	60	1009610.1	0.8	**521.0961**	4.502	0.9	ug/L	215	Standard
	Cu	65	102190.2	1.2	**52.9646**	0.037	0.1	ug/L	104	Standard
	Zn	66	1265218.7	0.3	**1119.1475**	11.121	1.0	ug/L	363	Standard
>	**Ge**	72	512284.2	1.2				ug/L	495142	Standard
	As	75	65072.9	0.3	**56.2148**	0.511	0.9	ug/L	-50	Standard
	Se	82	7179.2	1.0	**63.0878**	0.484	0.8	ug/L	14	Standard
	Se-1	77	4836.8	0.7	**63.3901**	0.749	1.2	ug/L	42	Standard
>	**Ga**	71	3398.7	7.2				mg/L	42	Standard
	Rb	85	158551.9	0.5				ug/L	17	Standard
	Y	89	3079691.2	1.3				ug/L	417098	Standard
>	**Rh**	103	243.3	6.3				ug/L	13	Standard
	Mo	98	610.5	4.6	**0.2160**	0.008	3.7	ug/L	11	Standard
	Ag	107	329069.5	0.6	**48.8691**	0.577	1.2	ug/L	68	Standard
	Cd	111	109320.3	0.2	**50.6101**	0.459	0.9	mg/L	2	Standard
	Cd	114	265491.9	0.4	**50.1152**	0.608	1.2	ug/L	12	Standard
>	**In**	115	606831.5	1.0				ug/L	587441	Standard
	Sn	118	1098.4	5.9	**0.0874**	0.012	14.0	ug/L	497	Standard
	Sb	123	240923.7	1.8	**49.9558**	1.171	2.3	ug/L	78	Standard
	Ba	135	113288.6	0.8	**53.5705**	0.782	1.5	ug/L	24	Standard
	Ce	140	2588807.2	1.1				ug/L	25	Standard
>	**Tb**	159	1203377.0	1.6				ug/L	939701	Standard
	Ho	165	210409.6	1.8				ug/L	10	Standard
	Tl	203	437720.2	1.4	**52.3305**	0.744	1.4	ug/L	4	Standard
	Tl	205	404409.4	0.6	**55.7069**	0.222	0.4	ug/L	2	Standard
	Pb	206	295264.9	0.3	**54.4023**	0.212	0.4	ug/L	240	Standard
	Pb	207	250626.5	0.1	**51.9726**	0.335	0.6	ug/L	198	Standard
	Pb	208	1033784.3	0.4	**54.2614**	0.547	1.0	ug/L	883	Standard
	U	238	423217.8	0.6	**60.5919**	0.666	1.1	ug/L	3	Standard
>	**Bi**	209	531687.2	0.6				ug/L	523984	Standard

Sample ID: L1606124701PS WG574688-01
Report Date/Time: Wednesday, June 29, 2016 16:16:41
Page 1

Approved: July 04, 2016

Na	23	31.7	32.9	**235.3667**	80.946	34.4	mg/L	0	Standard
Mg	24	88559.7	0.9	**73.3333**	1.498	2.0	mg/L	37	Standard
K	39	1311.7	6.5	**1.7460**	0.092	5.3	mg/L	38	Standard
Ca	43	3437.1	1.4	**136.8797**	3.742	2.7	mg/L	10	Standard
Fe	54	183063.9	0.7	**99.5077**	1.370	1.4	mg/L	179	Standard
Fe	57	49900.4	2.7	**107.8438**	4.358	4.0	mg/L	335	Standard
Sc-1	45	95075.8	2.0				mg/L	61029	Standard
Cl	35	63889.8	1.2				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	26186.0	7.9				ug/L	1547	Standard
P	31	52633.0	2.3				ug/L	15286	Standard
S	34	1903.5	2.7				ug/L	1572	Standard
Sr	88	1258.4	8.1				ug/L	73	Standard
C	12	153.3	16.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	329580.7	1.9				mg/L	6	Standard
Ho-1	165	210409.6	1.8				mg/L	10	Standard
Er	166	182077.6	1.3				mg/L	13	Standard
I	127	87547.6	5.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		222.720	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		103.462	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606124701PS WG574688-01
Report Date/Time: Wednesday, June 29, 2016 16:16:41
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.301
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.470
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701PS WG574688-01
Report Date/Time: Wednesday, June 29, 2016 16:16:41
Page 3

Approved: July 04, 2016

Co 59 Upper, S, EEE	Co	59
Ni 60 Upper, S, EEE	Ni	60
Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701SDL WG574688-02
Sample Date/Time: Wednesday, June 29, 2016 16:17:36
Number of Replicates: 3
Autosampler Position: 214
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	115043.6	2.7				ug/L	74699	Standard
	Be	9	2510.2	6.0	**1.8579**	0.064	3.4	ug/L	8	Standard
	Al	27	26603666.2	1.3	**117.5773**	3.118	2.7	ug/L	1493	Standard
	Sc	45	70520.4	2.0				ug/L	61029	Standard
	Ti	47	2655.2	2.4	**6.2496**	0.159	2.5	ug/L	35	Standard
	V	51	13651.6	1.3	**1.4451**	0.028	2.0	ug/L	1435	Standard
	Cr	52	32029.5	0.6	**2.9000**	0.113	3.9	ug/L	7635	Standard
	Cr	53	4062.2	6.0	**3.6565**	0.147	4.0	ug/L	363	Standard
	Mn	55	11326454.5	1.3	**1476.9710**	52.700	3.6	ug/L	747	Standard
	Co	59	315829.2	0.8	**43.3845**	1.293	3.0	ug/L	234	Standard
	Ni	60	206220.6	0.5	**99.8677**	2.554	2.6	ug/L	215	Standard
	Cu	65	3227.7	3.3	**1.4714**	0.087	5.9	ug/L	104	Standard
	Zn	66	276031.8	0.8	**228.9686**	6.765	3.0	ug/L	363	Standard
>	**Ge**	72	545622.5	2.3				ug/L	495142	Standard
	As	75	529.0	5.1	**0.4886**	0.031	6.3	ug/L	-50	Standard
	Se	82	109.8	7.8	**0.7692**	0.061	7.9	ug/L	14	Standard
	Se-1	77	127.7	10.6	**0.9669**	0.203	21.0	ug/L	42	Standard
>	**Ga**	71	896.7	4.3				mg/L	42	Standard
	Rb	85	33246.4	0.8				ug/L	17	Standard
	Y	89	1035711.2	2.7				ug/L	417098	Standard
>	**Rh**	103	68.3	33.0				ug/L	13	Standard
	Mo	98	137.2	4.4	**0.0416**	0.001	3.0	ug/L	11	Standard
	Ag	107	193.7	13.0	**0.0144**	0.003	20.5	ug/L	68	Standard
	Cd	111	922.1	1.5	**0.3978**	0.011	2.7	mg/L	2	Standard
	Cd	114	2202.4	8.0	**0.3929**	0.032	8.2	ug/L	12	Standard
>	**In**	115	638835.9	2.4				ug/L	587441	Standard
	Sn	118	601.7	4.7	**-0.0000**	0.004	76937.0	ug/L	497	Standard
	Sb	123	1960.6	22.3	**0.3834**	0.095	24.7	ug/L	78	Standard
	Ba	135	3410.7	1.0	**1.5171**	0.049	3.2	ug/L	24	Standard
	Ce	140	553030.9	1.0				ug/L	25	Standard
>	**Tb**	159	1064988.1	1.7				ug/L	939701	Standard
	Ho	165	44299.7	1.5				ug/L	10	Standard
	Tl	203	596.3	8.1	**0.0662**	0.006	8.4	ug/L	4	Standard
	Tl	205	613.3	10.1	**0.0808**	0.008	10.1	ug/L	2	Standard
	Pb	206	1759.1	0.5	**0.2383**	0.002	0.8	ug/L	240	Standard
	Pb	207	1407.7	5.3	**0.2088**	0.015	7.2	ug/L	198	Standard
	Pb	208	5898.4	2.6	**0.2216**	0.008	3.7	ug/L	883	Standard
	U	238	5357.3	1.2	**0.7271**	0.010	1.3	ug/L	3	Standard
>	**Bi**	209	557502.4	0.2				ug/L	523984	Standard

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:19:53
Page 1

Approved: July 04, 2016

Na	23	11.7	89.2	**117.5100**	105.501	89.8	mg/L	0	Standard
Mg	24	18105.8	1.6	**20.1624**	0.144	0.7	mg/L	37	Standard
K	39	321.7	3.6	**0.5016**	0.035	6.9	mg/L	38	Standard
Ca	43	733.4	8.9	**37.5359**	2.964	7.9	mg/L	10	Standard
Fe	54	38163.2	2.6	**27.8607**	0.186	0.7	mg/L	179	Standard
Fe	57	10221.8	1.3	**28.9959**	0.316	1.1	mg/L	335	Standard
> **Sc-1**	45	70520.4	2.0				mg/L	61029	Standard
Cl	35	31180.1	1.1				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	7325.1	3.6				ug/L	1547	Standard
P	31	24873.7	0.5				ug/L	15286	Standard
S	34	1646.8	8.5				ug/L	1572	Standard
Sr	88	296.7	11.0				ug/L	73	Standard
C	12	96.7	53.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	69181.7	2.0				mg/L	6	Standard
Ho-1	165	44299.7	1.5				mg/L	10	Standard
Er	166	39063.6	1.3				mg/L	13	Standard
I	127	21375.1	3.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		154.010	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.195	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:19:53
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.749
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.397
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:19:53
Page 3

Approved: July 04, 2016

Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 16:20:49
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	104492.8	3.9				ug/L	74699	Standard
	Be	9	63869.3	1.0	**51.9084**	1.516	2.9	ug/L	8	Standard
	Al	27	9148865.9	2.6	**44.5067**	0.724	1.6	ug/L	1493	Standard
	Sc	45	68447.7	3.2				ug/L	61029	Standard
	Ti	47	43923.6	1.6	**106.2047**	1.334	1.3	ug/L	35	Standard
	V	51	454775.2	0.5	**55.0469**	0.468	0.8	ug/L	1435	Standard
	Cr	52	431898.3	0.6	**52.4410**	0.316	0.6	ug/L	7635	Standard
	Cr	53	53343.8	2.8	**52.6203**	1.459	2.8	ug/L	363	Standard
	Mn	55	427711.0	4.2	**56.2006**	1.985	3.5	ug/L	747	Standard
	Co	59	425402.4	1.1	**59.0619**	0.614	1.0	ug/L	234	Standard
	Ni	60	106741.4	0.9	**52.1773**	0.185	0.4	ug/L	215	Standard
	Cu	65	106457.4	0.4	**52.3703**	0.176	0.3	ug/L	104	Standard
	Zn	66	63607.9	0.7	**52.9682**	0.530	1.0	ug/L	363	Standard
>	**Ge**	72	539725.1	0.7				ug/L	495142	Standard
	As	75	63945.9	1.5	**52.4378**	1.123	2.1	ug/L	-50	Standard
	Se	82	6584.0	0.6	**54.8988**	0.684	1.2	ug/L	14	Standard
	Se-1	77	4386.3	1.4	**54.4772**	1.111	2.0	ug/L	42	Standard
>	**Ga**	71	80.0	16.5				mg/L	42	Standard
	Rb	85	970.0	8.8				ug/L	17	Standard
	Y	89	494238.2	2.5				ug/L	417098	Standard
>	**Rh**	103	45.0	11.1				ug/L	13	Standard
	Mo	98	317512.9	0.8	**102.4795**	1.834	1.8	ug/L	11	Standard
	Ag	107	378789.7	0.9	**49.9860**	0.591	1.2	ug/L	68	Standard
	Cd	111	119686.6	0.4	**49.2403**	0.865	1.8	mg/L	2	Standard
	Cd	114	291890.5	1.4	**48.9669**	1.322	2.7	ug/L	12	Standard
>	**In**	115	682927.0	1.5				ug/L	587441	Standard
	Sn	118	338721.1	0.8	**49.7884**	0.723	1.5	ug/L	497	Standard
	Sb	123	263759.6	1.2	**48.5931**	0.331	0.7	ug/L	78	Standard
	Ba	135	111614.3	0.7	**46.9010**	0.972	2.1	ug/L	24	Standard
	Ce	140	1265.1	98.5				ug/L	25	Standard
>	**Tb**	159	1064058.1	1.0				ug/L	939701	Standard
	Ho	165	88.3	85.2				ug/L	10	Standard
	Tl	203	475263.7	0.7	**51.4258**	1.211	2.4	ug/L	4	Standard
	Tl	205	438160.3	0.4	**54.6297**	1.400	2.6	ug/L	2	Standard
	Pb	206	306629.0	0.6	**51.1326**	1.396	2.7	ug/L	240	Standard
	Pb	207	272793.9	0.3	**51.2040**	1.495	2.9	ug/L	198	Standard
	Pb	208	1100211.2	0.6	**52.2716**	1.709	3.3	ug/L	883	Standard
	U	238	414469.5	0.5	**53.7134**	1.693	3.2	ug/L	3	Standard
>	**Bi**	209	587705.0	2.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 16:23:06
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**35.5299**	61.531	173.2	mg/L	0	Standard
Mg	24	4809.1	8.0	**5.4678**	0.353	6.5	mg/L	37	Standard
K	39	2090.1	1.6	**4.0082**	0.167	4.2	mg/L	38	Standard
Ca	43	111.7	10.3	**3.7837**	0.862	22.8	mg/L	10	Standard
Fe	54	7082.4	2.6	**5.2191**	0.217	4.1	mg/L	179	Standard
Fe	57	2040.1	4.3	**5.1108**	0.267	5.2	mg/L	335	Standard
Sc-1	45	68447.7	3.2				mg/L	61029	Standard
Cl	35	25830.7	1.4				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	2190.2	6.5				ug/L	1547	Standard
P	31	18406.2	0.9				ug/L	15286	Standard
S	34	1965.1	6.7				ug/L	1572	Standard
Sr	88	71.7	21.3				ug/L	73	Standard
C	12	103.3	29.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	152.5	102.6				mg/L	6	Standard
Ho-1	165	88.3	85.2				mg/L	10	Standard
Er	166	86.7	74.2				mg/L	13	Standard
I	127	2766.9	21.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	103.817		
Al	27	89.013		
Sc	45			
Ti	47	106.205		
V	51	110.094		
Cr	52	104.882		
Cr	53			
Mn	55	112.401		
Co	59	118.124		
Ni	60	104.355		
Cu	65	104.741		
Zn	66	105.936		
Ge	72		109.004	
As	75	104.876		
Se	82	109.798		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 16:23:06
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.479		
Ag	107	99.972		
Cd	111	98.481		
Cd	114			
In	115		116.255	
Sn	118	99.577		
Sb	123	97.186		
Ba	135	93.802		
Ce	140			
Tb	159			
Ho	165			
Tl	203	102.852		
Tl	205			
Pb	206			
Pb	207			
Pb	208	104.543		
U	238	107.427		
Bi	209		112.161	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	V	51	
QC Std 6	Mn	55	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 16:23:06
Page 3

Approved: July 04, 2016

QC Std 6	Co	59

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 16:24:01
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100067.7	1.3				ug/L	74699	Standard
	Be	9	36.7	7.9	**0.0387**	0.002	5.2	ug/L	8	Standard
	Al	27	10153.6	30.4	**0.0381**	0.015	39.5	ug/L	1493	Standard
	Sc	45	67434.8	2.7				ug/L	61029	Standard
	Ti	47	38.3	21.7	**-0.0212**	0.020	92.7	ug/L	35	Standard
	V	51	1502.7	1.2	**-0.0231**	0.006	25.9	ug/L	1435	Standard
	Cr	52	7451.2	2.4	**-0.1523**	0.012	7.6	ug/L	7635	Standard
	Cr	53	503.3	11.6	**0.1429**	0.063	44.4	ug/L	363	Standard
	Mn	55	3693.5	26.8	**0.3137**	0.134	42.8	ug/L	747	Standard
	Co	59	358.3	10.7	**0.0099**	0.006	62.4	ug/L	234	Standard
	Ni	60	209.7	4.9	**-0.0268**	0.006	22.7	ug/L	215	Standard
	Cu	65	198.7	7.2	**-0.0110**	0.009	80.7	ug/L	104	Standard
	Zn	66	532.3	5.7	**-0.0316**	0.034	106.3	ug/L	363	Standard
>	**Ge**	72	571298.3	2.3				ug/L	495142	Standard
	As	75	-12.4	42.9	**0.0499**	0.004	8.8	ug/L	-50	Standard
	Se	82	23.5	9.9	**0.0476**	0.021	43.8	ug/L	14	Standard
	Se-1	77	58.7	7.7	**0.0749**	0.069	91.8	ug/L	42	Standard
>	**Ga**	71	75.0	26.7				mg/L	42	Standard
	Rb	85	36.7	55.1				ug/L	17	Standard
	Y	89	500557.3	2.5				ug/L	417098	Standard
>	**Rh**	103	30.0	44.1				ug/L	13	Standard
	Mo	98	301.0	26.9	**0.0924**	0.026	27.9	ug/L	11	Standard
	Ag	107	165.0	21.0	**0.0091**	0.004	49.2	ug/L	68	Standard
	Cd	111	19.8	28.3	**0.0003**	0.002	699.6	mg/L	2	Standard
	Cd	114	46.5	85.8	**0.0058**	0.007	116.4	ug/L	12	Standard
>	**In**	115	675355.4	0.8				ug/L	587441	Standard
	Sn	118	876.7	18.6	**0.0359**	0.025	69.7	ug/L	497	Standard
	Sb	123	599.5	22.8	**0.1076**	0.026	24.3	ug/L	78	Standard
	Ba	135	42.7	23.9	**0.0021**	0.004	201.6	ug/L	24	Standard
	Ce	140	140.0	34.4				ug/L	25	Standard
>	**Tb**	159	1044716.9	1.2				ug/L	939701	Standard
	Ho	165	20.0	50.0				ug/L	10	Standard
	Tl	203	103.0	21.9	**0.0094**	0.003	27.0	ug/L	4	Standard
	Tl	205	70.0	58.5	**0.0089**	0.005	57.1	ug/L	2	Standard
	Pb	206	338.7	4.4	**-0.0148**	0.002	16.0	ug/L	240	Standard
	Pb	207	277.7	9.1	**-0.0179**	0.005	28.1	ug/L	198	Standard
	Pb	208	1213.0	9.6	**-0.0165**	0.006	36.1	ug/L	883	Standard
	U	238	62.0	33.9	**0.0036**	0.003	76.1	ug/L	3	Standard
>	**Bi**	209	588682.5	1.2				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 16:26:17
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.0278**	29.484	173.2	mg/L	0	Standard
Mg	24	56.7	20.4	**0.0014**	0.015	1120.4	mg/L	37	Standard
K	39	36.7	28.4	**-0.0407**	0.021	51.2	mg/L	38	Standard
Ca	43	26.7	28.6	**-1.0029**	0.415	41.4	mg/L	10	Standard
Fe	54	187.4	12.0	**0.0055**	0.017	304.1	mg/L	179	Standard
Fe	57	261.7	6.1	**-0.2681**	0.071	26.5	mg/L	335	Standard
Sc-1	45	67434.8	2.7				mg/L	61029	Standard
Cl	35	26219.3	0.7				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	2243.5	3.0				ug/L	1547	Standard
P	31	19277.3	3.9				ug/L	15286	Standard
S	34	2016.8	11.3				ug/L	1572	Standard
Sr	88	70.0	35.7				ug/L	73	Standard
C	12	120.0	25.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	49.4	87.7				mg/L	6	Standard
Ho-1	165	20.0	50.0				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	2125.2	22.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.381	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 16:26:17
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.966
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		112.347
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 16:26:17
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701SDL WG574688-02
Sample Date/Time: Wednesday, June 29, 2016 16:27:13
Number of Replicates: 3
Autosampler Position: 215
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	105717.5	0.5				ug/L	74699	Standard
	Be	9	553.3	1.4	**0.4517**	0.004	0.8	ug/L	8	Standard
	Al	27	5662006.5	2.5	**27.2120**	0.787	2.9	ug/L	1493	Standard
	Sc	45	72131.5	3.8				ug/L	61029	Standard
	Ti	47	604.0	7.1	**1.2783**	0.109	8.5	ug/L	35	Standard
	V	51	3941.0	1.4	**0.2585**	0.002	0.9	ug/L	1435	Standard
	Cr	52	12715.3	2.3	**0.4691**	0.023	4.9	ug/L	7635	Standard
	Cr	53	1340.1	9.1	**0.9313**	0.103	11.1	ug/L	363	Standard
	Mn	55	2452895.1	0.8	**306.4368**	3.099	1.0	ug/L	747	Standard
	Co	59	67057.1	0.2	**8.7998**	0.112	1.3	ug/L	234	Standard
	Ni	60	46732.3	0.4	**21.5970**	0.233	1.1	ug/L	215	Standard
	Cu	65	904.0	6.3	**0.3188**	0.021	6.7	ug/L	104	Standard
	Zn	66	62660.6	1.1	**49.4618**	0.122	0.2	ug/L	363	Standard
>	**Ge**	72	569030.6	1.4				ug/L	495142	Standard
	As	75	59.1	19.2	**0.1055**	0.009	8.4	ug/L	-50	Standard
	Se	82	34.4	3.8	**0.1350**	0.014	10.1	ug/L	14	Standard
	Se-1	77	82.7	11.5	**0.3621**	0.106	29.4	ug/L	42	Standard
>	**Ga**	71	288.3	19.2				mg/L	42	Standard
	Rb	85	7325.1	4.9				ug/L	17	Standard
	Y	89	616528.9	0.8				ug/L	417098	Standard
>	**Rh**	103	13.3	78.1				ug/L	13	Standard
	Mo	98	112.4	8.7	**0.0298**	0.003	9.5	ug/L	11	Standard
	Ag	107	122.3	20.4	**0.0029**	0.003	108.6	ug/L	68	Standard
	Cd	111	198.8	14.6	**0.0723**	0.012	16.2	mg/L	2	Standard
	Cd	114	516.3	5.9	**0.0829**	0.006	6.9	ug/L	12	Standard
>	**In**	115	696026.2	0.9				ug/L	587441	Standard
	Sn	118	760.0	4.7	**0.0151**	0.006	37.2	ug/L	497	Standard
	Sb	123	350.2	21.1	**0.0592**	0.014	23.4	ug/L	78	Standard
	Ba	135	775.0	6.3	**0.3035**	0.019	6.3	ug/L	24	Standard
	Ce	140	119180.2	0.3				ug/L	25	Standard
>	**Tb**	159	1103880.2	1.1				ug/L	939701	Standard
	Ho	165	9261.2	2.5				ug/L	10	Standard
	Tl	203	155.7	9.5	**0.0145**	0.001	9.9	ug/L	4	Standard
	Tl	205	156.7	24.4	**0.0191**	0.005	23.7	ug/L	2	Standard
	Pb	206	668.3	5.9	**0.0368**	0.005	14.8	ug/L	240	Standard
	Pb	207	530.0	3.8	**0.0264**	0.004	13.6	ug/L	198	Standard
	Pb	208	2281.7	2.6	**0.0309**	0.002	6.9	ug/L	883	Standard
	U	238	1160.7	2.8	**0.1411**	0.004	2.6	ug/L	3	Standard
>	**Bi**	209	607087.1	0.9				ug/L	523984	Standard

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:29:30
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.5008**	28.572	173.2	mg/L	0	Standard
Mg	24	3728.8	3.6	**4.0090**	0.135	3.4	mg/L	37	Standard
K	39	85.0	42.4	**0.0464**	0.073	156.7	mg/L	38	Standard
Ca	43	183.3	11.4	**7.2814**	1.162	16.0	mg/L	10	Standard
Fe	54	8713.3	2.8	**6.1146**	0.148	2.4	mg/L	179	Standard
Fe	57	2506.9	1.7	**6.1424**	0.321	5.2	mg/L	335	Standard
> **Sc-1**	45	72131.5	3.8				mg/L	61029	Standard
Cl	35	26208.6	1.1				ug/L	20195	Standard
Kr	83	2.7	21.7				ug/L	4	Standard
Br	81	3227.0	4.8				ug/L	1547	Standard
P	31	20363.7	0.5				ug/L	15286	Standard
S	34	1863.4	4.5				ug/L	1572	Standard
Sr	88	105.0	21.8				ug/L	73	Standard
C	12	56.7	36.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	14855.4	4.1				mg/L	6	Standard
Ho-1	165	9261.2	2.5				mg/L	10	Standard
Er	166	8048.8	4.7				mg/L	13	Standard
I	127	5796.1	1.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		141.525	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.923	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:29:30
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	118.484
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	115.860
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:29:30
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126504
Sample Date/Time: Wednesday, June 29, 2016 16:30:25
Number of Replicates: 3
Autosampler Position: 216
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101552.6	1.9				ug/L	74699	Standard
	Be	9	25.0	72.1	**0.0287**	0.016	54.3	ug/L	8	Standard
	Al	27	16250197.0	1.8	**81.3578**	2.887	3.5	ug/L	1493	Standard
	Sc	45	67044.7	1.9				ug/L	61029	Standard
	Ti	47	394.3	10.9	**0.8437**	0.094	11.2	ug/L	35	Standard
	V	51	39093.2	1.3	**4.5488**	0.169	3.7	ug/L	1435	Standard
	Cr	52	19068.3	2.3	**1.3348**	0.101	7.6	ug/L	7635	Standard
	Cr	53	5326.0	20.8	**4.9620**	1.165	23.5	ug/L	363	Standard
	Mn	55	6239.0	2.5	**0.6731**	0.027	4.0	ug/L	747	Standard
	Co	59	851.0	2.5	**0.0808**	0.004	4.5	ug/L	234	Standard
	Ni	60	1951.8	2.1	**0.8314**	0.029	3.5	ug/L	215	Standard
	Cu	65	1321.1	4.3	**0.5464**	0.014	2.5	ug/L	104	Standard
	Zn	66	5985.8	0.6	**4.5673**	0.118	2.6	ug/L	363	Standard
>	**Ge**	72	540512.2	2.5				ug/L	495142	Standard
	As	75	1446.2	5.5	**1.2419**	0.041	3.3	ug/L	-50	Standard
	Se	82	841.6	2.6	**6.8893**	0.246	3.6	ug/L	14	Standard
	Se-1	77	609.0	15.5	**7.0254**	1.243	17.7	ug/L	42	Standard
>	**Ga**	71	105.0	12.6				mg/L	42	Standard
	Rb	85	1541.7	7.2				ug/L	17	Standard
	Y	89	489017.5	2.4				ug/L	417098	Standard
>	**Rh**	103	70.0	18.9				ug/L	13	Standard
	Mo	98	501.8	4.3	**0.1636**	0.005	2.9	ug/L	11	Standard
	Ag	107	97.0	7.4	**0.0005**	0.001	200.8	ug/L	68	Standard
	Cd	111	26.4	23.0	**0.0035**	0.003	79.3	mg/L	2	Standard
	Cd	114	74.1	18.4	**0.0109**	0.002	21.9	ug/L	12	Standard
>	**In**	115	652860.7	1.6				ug/L	587441	Standard
	Sn	118	840.0	9.7	**0.0347**	0.012	35.4	ug/L	497	Standard
	Sb	123	289.8	10.9	**0.0517**	0.007	12.7	ug/L	78	Standard
	Ba	135	17710.3	0.6	**7.7715**	0.161	2.1	ug/L	24	Standard
	Ce	140	1525.1	0.6				ug/L	25	Standard
>	**Tb**	159	1031518.0	1.9				ug/L	939701	Standard
	Ho	165	43.3	24.0				ug/L	10	Standard
	Tl	203	163.3	10.6	**0.0174**	0.002	9.6	ug/L	4	Standard
	Tl	205	133.3	16.9	**0.0182**	0.003	18.7	ug/L	2	Standard
	Pb	206	441.3	2.3	**0.0087**	0.001	8.9	ug/L	240	Standard
	Pb	207	362.0	4.8	**0.0037**	0.003	70.2	ug/L	198	Standard
	Pb	208	1549.4	2.6	**0.0058**	0.001	11.8	ug/L	883	Standard
	U	238	23538.3	0.9	**3.3015**	0.086	2.6	ug/L	3	Standard
>	**Bi**	209	542188.4	2.1				ug/L	523984	Standard

Sample ID: L1606126504
Report Date/Time: Wednesday, June 29, 2016 16:32:42
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**35.7447**	61.903	173.2	mg/L	0	Standard
Mg	24	119069.4	0.9	**139.8814**	3.117	2.2	mg/L	37	Standard
K	39	143.3	11.2	**0.1739**	0.028	15.9	mg/L	38	Standard
Ca	43	463.3	1.6	**24.1242**	0.657	2.7	mg/L	10	Standard
Fe	54	197.6	4.4	**0.0142**	0.006	45.1	mg/L	179	Standard
Fe	57	435.0	13.1	**0.2738**	0.196	71.7	mg/L	335	Standard
> **Sc-1**	45	67044.7	1.9				mg/L	61029	Standard
Cl	35	31470.7	2.7				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	73323.9	4.5				ug/L	1547	Standard
P	31	23608.4	2.5				ug/L	15286	Standard
S	34	1800.1	6.6				ug/L	1572	Standard
Sr	88	293.3	6.9				ug/L	73	Standard
C	12	193.3	41.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	86.5	19.9				mg/L	6	Standard
Ho-1	165	43.3	24.0				mg/L	10	Standard
Er	166	73.3	34.3				mg/L	13	Standard
I	127	54762.3	5.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		135.949	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		109.163	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606126504
Report Date/Time: Wednesday, June 29, 2016 16:32:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.136
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.474
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126505
Sample Date/Time: Wednesday, June 29, 2016 16:33:36
Number of Replicates: 3
Autosampler Position: 217
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	99450.4	1.4				ug/L	74699	Standard
	Be	9	58.3	30.1	**0.0575**	0.016	27.1	ug/L	8	Standard
	Al	27	32118811.5	1.4	**164.1461**	0.108	0.1	ug/L	1493	Standard
	Sc	45	70749.8	1.3				ug/L	61029	Standard
	Ti	47	642.0	2.2	**1.4252**	0.030	2.1	ug/L	35	Standard
	V	51	8228.7	11.8	**0.7915**	0.119	15.1	ug/L	1435	Standard
	Cr	52	11308.2	2.8	**0.3584**	0.044	12.4	ug/L	7635	Standard
	Cr	53	3712.1	10.4	**3.3091**	0.391	11.8	ug/L	363	Standard
	Mn	55	3368829.0	1.1	**438.1329**	6.772	1.5	ug/L	747	Standard
	Co	59	3400.7	28.1	**0.4294**	0.132	30.8	ug/L	234	Standard
	Ni	60	4538.7	3.1	**2.0720**	0.077	3.7	ug/L	215	Standard
	Cu	65	1589.1	4.3	**0.6697**	0.036	5.3	ug/L	104	Standard
	Zn	66	6182.6	5.7	**4.6729**	0.311	6.7	ug/L	363	Standard
>	**Ge**	72	546665.5	0.5				ug/L	495142	Standard
	As	75	6156.5	0.5	**5.0379**	0.006	0.1	ug/L	-50	Standard
	Se	82	164.0	1.8	**1.2151**	0.026	2.2	ug/L	14	Standard
	Se-1	77	258.7	9.7	**2.5862**	0.322	12.5	ug/L	42	Standard
>	**Ga**	71	188.3	13.1				mg/L	42	Standard
	Rb	85	1378.4	10.6				ug/L	17	Standard
	Y	89	484009.6	2.1				ug/L	417098	Standard
>	**Rh**	103	118.3	6.5				ug/L	13	Standard
	Mo	98	286.3	31.1	**0.0925**	0.030	32.6	ug/L	11	Standard
	Ag	107	183.3	43.5	**0.0128**	0.011	86.7	ug/L	68	Standard
	Cd	111	248.2	17.8	**0.1006**	0.018	18.1	mg/L	2	Standard
	Cd	114	590.9	17.2	**0.1032**	0.017	16.3	ug/L	12	Standard
>	**In**	115	642069.8	1.9				ug/L	587441	Standard
	Sn	118	836.7	5.6	**0.0365**	0.010	26.8	ug/L	497	Standard
	Sb	123	288.3	29.8	**0.0524**	0.017	32.5	ug/L	78	Standard
	Ba	135	131296.8	1.0	**58.6819**	0.595	1.0	ug/L	24	Standard
	Ce	140	6513.1	10.5				ug/L	25	Standard
>	**Tb**	159	1012113.9	1.4				ug/L	939701	Standard
	Ho	165	216.7	7.1				ug/L	10	Standard
	Tl	203	185.3	61.4	**0.0202**	0.013	66.8	ug/L	4	Standard
	Tl	205	276.7	88.2	**0.0379**	0.033	87.8	ug/L	2	Standard
	Pb	206	891.4	31.8	**0.0916**	0.052	56.6	ug/L	240	Standard
	Pb	207	748.7	30.8	**0.0839**	0.048	56.6	ug/L	198	Standard
	Pb	208	3050.1	30.7	**0.0846**	0.049	57.6	ug/L	883	Standard
	U	238	21997.6	2.6	**3.1135**	0.091	2.9	ug/L	3	Standard
>	**Bi**	209	537104.2	0.4				ug/L	523984	Standard

Sample ID: L1606126505
Report Date/Time: Wednesday, June 29, 2016 16:35:53
Page 1

Approved: July 04, 2016

Na	23	13.3	57.3	**132.8617**	77.017	58.0	mg/L	0	Standard
Mg	24	104917.1	1.8	**116.7540**	0.597	0.5	mg/L	37	Standard
K	39	116.7	8.9	**0.1082**	0.018	16.4	mg/L	38	Standard
Ca	43	916.7	6.0	**47.4532**	3.610	7.6	mg/L	10	Standard
Fe	54	3013.9	2.4	**2.0666**	0.083	4.0	mg/L	179	Standard
Fe	57	1381.7	4.1	**2.9770**	0.162	5.4	mg/L	335	Standard
> **Sc-1**	45	70749.8	1.3				mg/L	61029	Standard
Cl	35	30215.4	1.0				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	35768.9	2.6				ug/L	1547	Standard
P	31	25429.7	2.5				ug/L	15286	Standard
S	34	2313.5	0.3				ug/L	1572	Standard
Sr	88	551.7	8.2				ug/L	73	Standard
C	12	433.3	25.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	3	Standard
Dy	164	279.9	12.3				mg/L	6	Standard
Ho-1	165	216.7	7.1				mg/L	10	Standard
Er	166	280.0	18.9				mg/L	13	Standard
I	127	794211.5	6.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		133.135	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.406	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126505
Report Date/Time: Wednesday, June 29, 2016 16:35:53
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	109.299
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.504
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606126505
Report Date/Time: Wednesday, June 29, 2016 16:35:53
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126506
Sample Date/Time: Wednesday, June 29, 2016 16:36:48
Number of Replicates: 3
Autosampler Position: 218
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100252.3	1.1				ug/L	74699	Standard
	Be	9	26.7	39.0	**0.0303**	0.009	29.8	ug/L	8	Standard
	Al	27	69444.0	4.9	**0.3387**	0.017	5.1	ug/L	1493	Standard
	Sc	45	71205.3	1.8				ug/L	61029	Standard
	Ti	47	113.0	20.9	**0.1414**	0.055	38.5	ug/L	35	Standard
	V	51	2901.2	2.9	**0.1267**	0.013	10.0	ug/L	1435	Standard
	Cr	52	23756.9	1.1	**1.6651**	0.003	0.2	ug/L	7635	Standard
	Cr	53	2710.2	4.1	**2.1308**	0.127	5.9	ug/L	363	Standard
	Mn	55	7460.8	10.3	**0.7511**	0.098	13.0	ug/L	747	Standard
	Co	59	456.0	8.6	**0.0207**	0.006	26.8	ug/L	234	Standard
	Ni	60	2497.2	2.2	**0.9949**	0.023	2.3	ug/L	215	Standard
	Cu	65	984.0	1.2	**0.3398**	0.007	2.0	ug/L	104	Standard
	Zn	66	8303.3	0.9	**5.9218**	0.058	1.0	ug/L	363	Standard
>	**Ge**	72	590298.4	1.1				ug/L	495142	Standard
	As	75	7.9	237.8	**0.0655**	0.014	21.6	ug/L	-50	Standard
	Se	82	27.7	14.2	**0.0742**	0.032	43.2	ug/L	14	Standard
	Se-1	77	72.3	15.2	**0.2091**	0.134	63.9	ug/L	42	Standard
>	**Ga**	71	91.7	20.7				mg/L	42	Standard
	Rb	85	308.3	4.1				ug/L	17	Standard
	Y	89	498995.2	1.7				ug/L	417098	Standard
>	**Rh**	103	21.7	48.0				ug/L	13	Standard
	Mo	98	716.7	9.5	**0.2259**	0.023	10.1	ug/L	11	Standard
	Ag	107	155.3	16.4	**0.0076**	0.003	44.6	ug/L	68	Standard
	Cd	111	44.4	12.9	**0.0104**	0.002	21.8	mg/L	2	Standard
	Cd	114	117.5	18.7	**0.0177**	0.004	21.5	ug/L	12	Standard
>	**In**	115	682064.0	0.6				ug/L	587441	Standard
	Sn	118	4689.1	2.4	**0.5966**	0.013	2.3	ug/L	497	Standard
	Sb	123	228.7	11.3	**0.0380**	0.005	12.5	ug/L	78	Standard
	Ba	135	1034.0	4.7	**0.4191**	0.020	4.8	ug/L	24	Standard
	Ce	140	486.7	14.8				ug/L	25	Standard
>	**Tb**	159	1056147.0	0.9				ug/L	939701	Standard
	Ho	165	8.3	69.3				ug/L	10	Standard
	Tl	203	43.7	26.0	**0.0029**	0.001	43.8	ug/L	4	Standard
	Tl	205	50.0	26.5	**0.0063**	0.002	26.5	ug/L	2	Standard
	Pb	206	665.7	3.1	**0.0376**	0.003	7.4	ug/L	240	Standard
	Pb	207	542.7	5.2	**0.0299**	0.006	20.6	ug/L	198	Standard
	Pb	208	2227.4	4.5	**0.0297**	0.006	20.4	ug/L	883	Standard
	U	238	38.7	20.9	**0.0005**	0.001	221.1	ug/L	3	Standard
>	**Bi**	209	599996.8	1.5				ug/L	523984	Standard

Sample ID: L1606126506
Report Date/Time: Wednesday, June 29, 2016 16:39:04
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	216.7	5.8	**0.1745**	0.010	5.5	mg/L	37	Standard
K	39	46.7	49.5	**-0.0257**	0.044	170.8	mg/L	38	Standard
Ca	43	30.0	16.7	**-0.8971**	0.290	32.3	mg/L	10	Standard
Fe	54	200.7	16.8	**0.0076**	0.025	326.7	mg/L	179	Standard
Fe	57	248.3	11.1	**-0.3516**	0.067	19.2	mg/L	335	Standard
> **Sc-1**	45	71205.3	1.8				mg/L	61029	Standard
Cl	35	31458.6	1.1				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	3110.3	10.8				ug/L	1547	Standard
P	31	21037.9	1.6				ug/L	15286	Standard
S	34	2263.5	3.0				ug/L	1572	Standard
Sr	88	58.3	27.6				ug/L	73	Standard
C	12	113.3	13.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	38.9	27.0				mg/L	6	Standard
Ho-1	165	8.3	69.3				mg/L	10	Standard
Er	166	23.3	65.5				mg/L	13	Standard
I	127	20210.2	61.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		134.209	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		119.218	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126506
Report Date/Time: Wednesday, June 29, 2016 16:39:04
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.108
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		114.507
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126601
Sample Date/Time: Wednesday, June 29, 2016 16:39:59
Number of Replicates: 3
Autosampler Position: 219
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101321.2	3.1				ug/L	74699	Standard
	Be	9	18.3	56.8	**0.0228**	0.008	36.0	ug/L	8	Standard
	Al	27	44431569.1	0.7	**222.9926**	5.415	2.4	ug/L	1493	Standard
	Sc	45	71310.8	2.2				ug/L	61029	Standard
	Ti	47	275.3	6.3	**0.5272**	0.047	8.9	ug/L	35	Standard
	V	51	27907.2	1.9	**3.0387**	0.030	1.0	ug/L	1435	Standard
	Cr	52	13580.1	0.8	**0.5803**	0.015	2.6	ug/L	7635	Standard
	Cr	53	2821.9	2.8	**2.3430**	0.081	3.4	ug/L	363	Standard
	Mn	55	153037.1	1.2	**19.0926**	0.078	0.4	ug/L	747	Standard
	Co	59	1012.0	1.5	**0.0969**	0.001	1.5	ug/L	234	Standard
	Ni	60	4370.3	2.4	**1.9194**	0.048	2.5	ug/L	215	Standard
	Cu	65	751.7	5.3	**0.2499**	0.020	8.0	ug/L	104	Standard
	Zn	66	3518.1	1.4	**2.3650**	0.050	2.1	ug/L	363	Standard
>	**Ge**	72	565652.8	1.1				ug/L	495142	Standard
	As	75	1269.0	4.3	**1.0512**	0.035	3.4	ug/L	-50	Standard
	Se	82	382.5	3.3	**2.9130**	0.125	4.3	ug/L	14	Standard
	Se-1	77	193.3	14.8	**1.6975**	0.369	21.7	ug/L	42	Standard
>	**Ga**	71	66.7	18.9				mg/L	42	Standard
	Rb	85	2323.5	7.7				ug/L	17	Standard
	Y	89	480038.2	1.2				ug/L	417098	Standard
>	**Rh**	103	150.0	3.3				ug/L	13	Standard
	Mo	98	638.7	4.4	**0.2084**	0.010	5.0	ug/L	11	Standard
	Ag	107	92.0	14.1	**-0.0003**	0.002	569.4	ug/L	68	Standard
	Cd	111	195.2	3.6	**0.0755**	0.003	3.5	mg/L	2	Standard
	Cd	114	469.8	19.5	**0.0797**	0.016	19.7	ug/L	12	Standard
>	**In**	115	657497.5	0.6				ug/L	587441	Standard
	Sn	118	753.4	8.5	**0.0205**	0.010	48.9	ug/L	497	Standard
	Sb	123	222.8	13.6	**0.0384**	0.006	15.3	ug/L	78	Standard
	Ba	135	100054.6	0.7	**43.6617**	0.503	1.2	ug/L	24	Standard
	Ce	140	173.3	6.0				ug/L	25	Standard
>	**Tb**	159	1045560.4	1.5				ug/L	939701	Standard
	Ho	165	40.0	25.0				ug/L	10	Standard
	Tl	203	129.3	1.6	**0.0127**	0.000	3.7	ug/L	4	Standard
	Tl	205	133.3	20.7	**0.0174**	0.003	20.0	ug/L	2	Standard
	Pb	206	412.3	3.7	**0.0001**	0.002	2456.7	ug/L	240	Standard
	Pb	207	343.7	8.9	**-0.0032**	0.005	165.7	ug/L	198	Standard
	Pb	208	1432.7	1.3	**-0.0035**	0.002	56.3	ug/L	883	Standard
	U	238	43636.7	0.7	**5.8505**	0.120	2.1	ug/L	3	Standard
>	**Bi**	209	567479.2	1.8				ug/L	523984	Standard

Sample ID: L1606126601
Report Date/Time: Wednesday, June 29, 2016 16:42:16
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**48.6438**	48.507	99.7	mg/L	0	Standard
Mg	24	35550.0	1.9	**39.2140**	0.699	1.8	mg/L	37	Standard
K	39	165.0	24.1	**0.1975**	0.071	36.0	mg/L	38	Standard
Ca	43	596.7	5.4	**29.7553**	2.186	7.3	mg/L	10	Standard
Fe	54	190.1	10.6	**-0.0004**	0.014	3403.7	mg/L	179	Standard
Fe	57	555.0	7.7	**0.5391**	0.093	17.2	mg/L	335	Standard
> **Sc-1**	45	71310.8	2.2				mg/L	61029	Standard
Cl	35	33374.1	0.9				ug/L	20195	Standard
Kr	83	2.0	132.3				ug/L	4	Standard
Br	81	77209.6	1.2				ug/L	1547	Standard
P	31	24828.7	1.5				ug/L	15286	Standard
S	34	2263.5	4.7				ug/L	1572	Standard
Sr	88	591.7	14.4				ug/L	73	Standard
C	12	253.3	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	40.0	25.0				mg/L	3	Standard
Dy	164	40.1	59.8				mg/L	6	Standard
Ho-1	165	40.0	25.0				mg/L	10	Standard
Er	166	66.7	37.7				mg/L	13	Standard
I	127	2058609.0	7.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		135.640	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.240	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126601
Report Date/Time: Wednesday, June 29, 2016 16:42:16
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.926
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.301
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126901
Sample Date/Time: Wednesday, June 29, 2016 16:43:10
Number of Replicates: 3
Autosampler Position: 220
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	97179.0	1.2				ug/L	74699	Standard
	Be	9	23.3	44.6	**0.0281**	0.009	33.4	ug/L	8	Standard
	Al	27	25054309.4	1.7	**131.0332**	1.604	1.2	ug/L	1493	Standard
	Sc	45	69363.5	2.7				ug/L	61029	Standard
	Ti	47	323.3	2.8	**0.6992**	0.021	2.9	ug/L	35	Standard
	V	51	12024.0	2.6	**1.3128**	0.037	2.9	ug/L	1435	Standard
	Cr	52	11840.6	1.6	**0.4900**	0.025	5.0	ug/L	7635	Standard
	Cr	53	2901.9	3.9	**2.6451**	0.119	4.5	ug/L	363	Standard
	Mn	55	490032.6	0.8	**66.5395**	0.816	1.2	ug/L	747	Standard
	Co	59	931.7	2.1	**0.0964**	0.003	3.3	ug/L	234	Standard
	Ni	60	2217.5	3.2	**0.9981**	0.030	3.0	ug/L	215	Standard
	Cu	65	599.0	3.0	**0.2013**	0.011	5.4	ug/L	104	Standard
	Zn	66	4806.1	1.2	**3.7141**	0.039	1.0	ug/L	363	Standard
>	**Ge**	72	522601.4	0.6				ug/L	495142	Standard
	As	75	2361.3	3.3	**2.0571**	0.073	3.5	ug/L	-50	Standard
	Se	82	593.9	4.2	**4.9896**	0.230	4.6	ug/L	14	Standard
	Se-1	77	179.7	2.8	**1.7086**	0.076	4.5	ug/L	42	Standard
>	**Ga**	71	105.0	9.5				mg/L	42	Standard
	Rb	85	2805.3	2.5				ug/L	17	Standard
	Y	89	474068.5	3.1				ug/L	417098	Standard
>	**Rh**	103	81.7	12.7				ug/L	13	Standard
	Mo	98	305.2	1.9	**0.0985**	0.004	4.4	ug/L	11	Standard
	Ag	107	99.7	12.7	**0.0010**	0.002	153.4	ug/L	68	Standard
	Cd	111	19.2	16.0	**0.0004**	0.001	278.0	mg/L	2	Standard
	Cd	114	49.5	42.7	**0.0067**	0.004	59.0	ug/L	12	Standard
>	**In**	115	645532.6	2.3				ug/L	587441	Standard
	Sn	118	888.4	9.9	**0.0435**	0.010	24.0	ug/L	497	Standard
	Sb	123	240.8	4.4	**0.0427**	0.002	4.3	ug/L	78	Standard
	Ba	135	155622.1	0.5	**69.2020**	1.756	2.5	ug/L	24	Standard
	Ce	140	928.4	7.9				ug/L	25	Standard
>	**Tb**	159	1004174.2	1.0				ug/L	939701	Standard
	Ho	165	23.3	32.7				ug/L	10	Standard
	Tl	203	115.7	10.5	**0.0121**	0.001	11.5	ug/L	4	Standard
	Tl	205	128.3	22.2	**0.0179**	0.004	22.0	ug/L	2	Standard
	Pb	206	395.0	0.9	**0.0016**	0.001	57.1	ug/L	240	Standard
	Pb	207	323.3	6.0	**-0.0029**	0.004	138.0	ug/L	198	Standard
	Pb	208	1415.7	2.0	**0.0002**	0.001	432.7	ug/L	883	Standard
	U	238	7265.7	2.0	**1.0348**	0.023	2.2	ug/L	3	Standard
>	**Bi**	209	532226.3	0.6				ug/L	523984	Standard

Sample ID: L1606126901
Report Date/Time: Wednesday, June 29, 2016 16:45:27
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.4177**	30.160	173.2	mg/L	0	Standard
Mg	24	119202.1	1.4	**135.3768**	3.695	2.7	mg/L	37	Standard
K	39	220.0	8.2	**0.3143**	0.042	13.4	mg/L	38	Standard
Ca	43	398.3	4.4	**19.6130**	0.698	3.6	mg/L	10	Standard
Fe	54	252.1	18.1	**0.0493**	0.029	59.6	mg/L	179	Standard
Fe	57	418.3	9.7	**0.1785**	0.138	77.4	mg/L	335	Standard
Sc-1	45	69363.5	2.7				mg/L	61029	Standard
Cl	35	32454.7	1.8				ug/L	20195	Standard
Kr	83	1.0	173.2				ug/L	4	Standard
Br	81	137582.4	3.7				ug/L	1547	Standard
P	31	24828.7	1.5				ug/L	15286	Standard
S	34	2306.8	3.7				ug/L	1572	Standard
Sr	88	291.7	6.0				ug/L	73	Standard
C	12	323.3	19.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	15.2	73.1				mg/L	6	Standard
Ho-1	165	23.3	32.7				mg/L	10	Standard
Er	166	30.0	33.3				mg/L	13	Standard
I	127	310633.7	2.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		130.094	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		105.546	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606126901
Report Date/Time: Wednesday, June 29, 2016 16:45:27
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.889
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.573
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127201
Sample Date/Time: Wednesday, June 29, 2016 16:46:22
Number of Replicates: 3
Autosampler Position: 221
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94337.9	2.3				ug/L	74699	Standard
	Be	9	23.3	53.9	**0.0288**	0.012	40.8	ug/L	8	Standard
	Al	27	2978340.1	2.2	**16.0395**	0.522	3.3	ug/L	1493	Standard
	Sc	45	67137.0	4.7				ug/L	61029	Standard
	Ti	47	236.7	18.3	**0.4520**	0.099	21.9	ug/L	35	Standard
	V	51	33734.3	0.8	**3.8201**	0.030	0.8	ug/L	1435	Standard
	Cr	52	13200.4	3.0	**0.5779**	0.062	10.7	ug/L	7635	Standard
	Cr	53	1396.7	5.5	**1.0289**	0.077	7.5	ug/L	363	Standard
	Mn	55	5810.0	55.0	**0.6034**	0.416	68.9	ug/L	747	Standard
	Co	59	597.0	23.2	**0.0441**	0.019	43.4	ug/L	234	Standard
	Ni	60	839.7	12.2	**0.2798**	0.055	19.7	ug/L	215	Standard
	Cu	65	499.7	4.1	**0.1378**	0.013	9.3	ug/L	104	Standard
	Zn	66	1881.8	10.3	**1.0958**	0.181	16.5	ug/L	363	Standard
>	**Ge**	72	550779.1	1.5				ug/L	495142	Standard
	As	75	823.5	3.7	**0.7205**	0.021	2.9	ug/L	-50	Standard
	Se	82	534.3	3.8	**4.2403**	0.211	5.0	ug/L	14	Standard
	Se-1	77	394.7	5.2	**4.2338**	0.183	4.3	ug/L	42	Standard
>	**Ga**	71	76.7	13.6				mg/L	42	Standard
	Rb	85	1120.0	6.4				ug/L	17	Standard
	Y	89	478775.5	2.0				ug/L	417098	Standard
>	**Rh**	103	33.3	52.7				ug/L	13	Standard
	Mo	98	1510.8	4.0	**0.5152**	0.028	5.4	ug/L	11	Standard
	Ag	107	155.0	42.6	**0.0091**	0.010	106.9	ug/L	68	Standard
	Cd	111	66.6	64.6	**0.0216**	0.019	89.7	mg/L	2	Standard
	Cd	114	188.5	50.4	**0.0319**	0.018	55.0	ug/L	12	Standard
>	**In**	115	639482.4	2.0				ug/L	587441	Standard
	Sn	118	738.4	13.9	**0.0215**	0.017	79.5	ug/L	497	Standard
	Sb	123	301.8	22.4	**0.0554**	0.014	26.0	ug/L	78	Standard
	Ba	135	10964.6	1.7	**4.9054**	0.035	0.7	ug/L	24	Standard
	Ce	140	728.4	65.5				ug/L	25	Standard
>	**Tb**	159	995838.4	2.1				ug/L	939701	Standard
	Ho	165	55.0	78.7				ug/L	10	Standard
	Tl	203	165.3	64.1	**0.0175**	0.012	69.7	ug/L	4	Standard
	Tl	205	88.3	34.1	**0.0121**	0.004	34.3	ug/L	2	Standard
	Pb	206	371.3	17.5	**-0.0043**	0.012	275.3	ug/L	240	Standard
	Pb	207	297.3	13.9	**-0.0098**	0.008	82.9	ug/L	198	Standard
	Pb	208	1253.7	10.6	**-0.0098**	0.007	68.1	ug/L	883	Standard
	U	238	7715.3	1.4	**1.0741**	0.027	2.5	ug/L	3	Standard
>	**Bi**	209	544657.1	1.1				ug/L	523984	Standard

Sample ID: L1606127201
Report Date/Time: Wednesday, June 29, 2016 16:48:39
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**87.3928**	30.523	34.9	mg/L	0	Standard
Mg	24	56818.0	2.0	**66.6625**	1.829	2.7	mg/L	37	Standard
K	39	118.3	28.8	**0.1220**	0.059	48.6	mg/L	38	Standard
Ca	43	173.3	14.2	**7.4463**	1.537	20.6	mg/L	10	Standard
Fe	54	200.1	8.9	**0.0164**	0.019	116.3	mg/L	179	Standard
Fe	57	341.7	12.8	**-0.0156**	0.156	997.8	mg/L	335	Standard
> **Sc-1**	45	67137.0	4.7				mg/L	61029	Standard
Cl	35	31870.8	2.6				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	13232.4	3.4				ug/L	1547	Standard
P	31	22137.8	2.1				ug/L	15286	Standard
S	34	2160.2	8.0				ug/L	1572	Standard
Sr	88	146.7	24.2				ug/L	73	Standard
C	12	170.0	10.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	64.3	100.4				mg/L	6	Standard
Ho-1	165	55.0	78.7				mg/L	10	Standard
Er	166	50.0	60.0				mg/L	13	Standard
I	127	62898.8	4.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		126.291	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.237	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127201
Report Date/Time: Wednesday, June 29, 2016 16:48:39
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.859
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.945
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127202
Sample Date/Time: Wednesday, June 29, 2016 16:49:33
Number of Replicates: 3
Autosampler Position: 222
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98788.0	4.1				ug/L	74699	Standard
	Be	9	13.3	57.3	**0.0192**	0.007	36.1	ug/L	8	Standard
	Al	27	43121483.9	2.2	**221.9923**	5.444	2.5	ug/L	1493	Standard
	Sc	45	66930.9	3.1				ug/L	61029	Standard
	Ti	47	1039.7	26.4	**2.4141**	0.661	27.4	ug/L	35	Standard
	V	51	49957.2	0.9	**5.8910**	0.067	1.1	ug/L	1435	Standard
	Cr	52	18344.1	1.8	**1.2538**	0.038	3.0	ug/L	7635	Standard
	Cr	53	24960.6	4.5	**24.5202**	1.080	4.4	ug/L	363	Standard
	Mn	55	17873.9	4.1	**2.2143**	0.103	4.7	ug/L	747	Standard
	Co	59	1405.4	6.7	**0.1586**	0.013	8.4	ug/L	234	Standard
	Ni	60	3876.2	1.9	**1.7810**	0.032	1.8	ug/L	215	Standard
	Cu	65	1407.4	2.7	**0.5921**	0.020	3.4	ug/L	104	Standard
	Zn	66	4623.4	0.6	**3.4401**	0.014	0.4	ug/L	363	Standard
>	**Ge**	72	538103.3	0.3				ug/L	495142	Standard
	As	75	2665.5	0.4	**2.2493**	0.008	0.3	ug/L	-50	Standard
	Se	82	859.9	5.4	**7.0717**	0.401	5.7	ug/L	14	Standard
	Se-1	77	1900.1	5.6	**23.3152**	1.302	5.6	ug/L	42	Standard
>	**Ga**	71	205.0	14.8				mg/L	42	Standard
	Rb	85	12111.5	1.9				ug/L	17	Standard
	Y	89	447538.5	3.5				ug/L	417098	Standard
>	**Rh**	103	186.7	5.6				ug/L	13	Standard
	Mo	98	4087.3	1.0	**1.4942**	0.026	1.7	ug/L	11	Standard
	Ag	107	258.7	10.6	**0.0259**	0.004	14.5	ug/L	68	Standard
	Cd	111	70.2	15.3	**0.0250**	0.005	21.6	mg/L	2	Standard
	Cd	114	224.5	9.2	**0.0407**	0.004	8.7	ug/L	12	Standard
>	**In**	115	600654.0	1.1				ug/L	587441	Standard
	Sn	118	765.0	14.6	**0.0335**	0.020	60.1	ug/L	497	Standard
	Sb	123	166.1	4.9	**0.0306**	0.002	5.1	ug/L	78	Standard
	Ba	135	52090.1	1.0	**24.8744**	0.038	0.2	ug/L	24	Standard
	Ce	140	4474.0	3.6				ug/L	25	Standard
>	**Tb**	159	974974.5	0.9				ug/L	939701	Standard
	Ho	165	80.0	27.2				ug/L	10	Standard
	Tl	203	82.0	10.0	**0.0082**	0.001	13.1	ug/L	4	Standard
	Tl	205	63.3	43.5	**0.0090**	0.004	42.2	ug/L	2	Standard
	Pb	206	1123.7	3.1	**0.1383**	0.006	4.7	ug/L	240	Standard
	Pb	207	914.0	2.3	**0.1218**	0.004	3.1	ug/L	198	Standard
	Pb	208	3779.9	2.5	**0.1267**	0.006	5.0	ug/L	883	Standard
	U	238	23648.8	1.1	**3.4170**	0.057	1.7	ug/L	3	Standard
>	**Bi**	209	526202.0	0.7				ug/L	523984	Standard

Sample ID: L1606127202
Report Date/Time: Wednesday, June 29, 2016 16:51:50
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**53.6294**	53.947	100.6	mg/L	0	Standard
Mg	24	113104.0	1.5	**133.1465**	4.642	3.5	mg/L	37	Standard
K	39	478.3	11.2	**0.8485**	0.082	9.7	mg/L	38	Standard
Ca	43	695.0	6.9	**37.5535**	3.460	9.2	mg/L	10	Standard
Fe	54	250.5	4.6	**0.0553**	0.004	6.4	mg/L	179	Standard
Fe	57	631.7	3.6	**0.8833**	0.015	1.7	mg/L	335	Standard
> **Sc-1**	45	66930.9	3.1				mg/L	61029	Standard
Cl	35	32927.1	2.7				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	41968.1	3.2				ug/L	1547	Standard
P	31	23413.1	2.8				ug/L	15286	Standard
S	34	2041.8	4.5				ug/L	1572	Standard
Sr	88	750.0	13.9				ug/L	73	Standard
C	12	126.7	19.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	103.6	24.2				mg/L	6	Standard
Ho-1	165	80.0	27.2				mg/L	10	Standard
Er	166	63.3	32.9				mg/L	13	Standard
I	127	36619.2	2.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.248	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.676	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127202
Report Date/Time: Wednesday, June 29, 2016 16:51:50
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.249
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.423
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127901
Sample Date/Time: Wednesday, June 29, 2016 16:52:45
Number of Replicates: 3
Autosampler Position: 223
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94669.8	0.8				ug/L	74699	Standard
	Be	9	153.3	14.7	**0.1451**	0.021	14.4	ug/L	8	Standard
	Al	27	36828943.8	1.6	**197.7220**	2.539	1.3	ug/L	1493	Standard
	Sc	45	72494.8	2.9				ug/L	61029	Standard
	Ti	47	9823.2	1.5	**22.6070**	0.583	2.6	ug/L	35	Standard
	V	51	40160.9	1.2	**4.4642**	0.016	0.4	ug/L	1435	Standard
	Cr	52	82051.3	1.3	**8.6794**	0.190	2.2	ug/L	7635	Standard
	Cr	53	10588.7	4.0	**9.7108**	0.425	4.4	ug/L	363	Standard
	Mn	55	3156665.4	1.9	**397.2480**	11.605	2.9	ug/L	747	Standard
	Co	59	13202.1	1.7	**1.7150**	0.048	2.8	ug/L	234	Standard
	Ni	60	14584.3	1.0	**6.7029**	0.116	1.7	ug/L	215	Standard
	Cu	65	6796.9	2.1	**3.0972**	0.099	3.2	ug/L	104	Standard
	Zn	66	29601.9	0.7	**23.2962**	0.407	1.7	ug/L	363	Standard
>	**Ge**	72	565029.1	1.1				ug/L	495142	Standard
	As	75	1169.4	2.0	**0.9746**	0.024	2.5	ug/L	-50	Standard
	Se	82	54.4	2.8	**0.2964**	0.014	4.6	ug/L	14	Standard
	Se-1	77	95.3	8.8	**0.5209**	0.088	16.8	ug/L	42	Standard
>	**Ga**	71	2745.3	10.7				mg/L	42	Standard
	Rb	85	40779.8	2.2				ug/L	17	Standard
	Y	89	501434.2	0.6				ug/L	417098	Standard
>	**Rh**	103	116.7	40.7				ug/L	13	Standard
	Mo	98	6510.4	0.7	**2.1843**	0.030	1.4	ug/L	11	Standard
	Ag	107	146.0	1.2	**0.0072**	0.000	3.6	ug/L	68	Standard
	Cd	111	54.1	19.4	**0.0153**	0.004	28.2	mg/L	2	Standard
	Cd	114	169.8	20.2	**0.0276**	0.006	22.4	ug/L	12	Standard
>	**In**	115	655252.5	1.2				ug/L	587441	Standard
	Sn	118	1435.1	4.4	**0.1255**	0.007	5.6	ug/L	497	Standard
	Sb	123	419.8	7.1	**0.0764**	0.005	6.8	ug/L	78	Standard
	Ba	135	167537.6	1.4	**73.3690**	0.619	0.8	ug/L	24	Standard
	Ce	140	144147.5	0.8				ug/L	25	Standard
>	**Tb**	159	1041877.9	1.3				ug/L	939701	Standard
	Ho	165	2731.9	2.9				ug/L	10	Standard
	Tl	203	237.3	5.5	**0.0244**	0.001	5.3	ug/L	4	Standard
	Tl	205	228.3	8.3	**0.0292**	0.002	8.4	ug/L	2	Standard
	Pb	206	7357.1	0.9	**1.1818**	0.018	1.5	ug/L	240	Standard
	Pb	207	5920.5	2.7	**1.0648**	0.026	2.5	ug/L	198	Standard
	Pb	208	25273.6	1.1	**1.1521**	0.016	1.4	ug/L	883	Standard
	U	238	16614.4	0.8	**2.1915**	0.022	1.0	ug/L	3	Standard
>	**Bi**	209	575969.2	0.6				ug/L	523984	Standard

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:55:02
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**32.3968**	28.087	86.7	mg/L	0	Standard
Mg	24	3568.8	4.7	**3.8119**	0.091	2.4	mg/L	37	Standard
K	39	285.0	24.6	**0.4167**	0.132	31.6	mg/L	38	Standard
Ca	43	908.4	6.1	**45.7783**	2.516	5.5	mg/L	10	Standard
Fe	54	3605.6	1.3	**2.4364**	0.072	2.9	mg/L	179	Standard
Fe	57	1446.7	0.5	**3.0674**	0.109	3.5	mg/L	335	Standard
Sc-1	45	72494.8	2.9				mg/L	61029	Standard
Cl	35	31293.6	1.2				ug/L	20195	Standard
Kr	83	2.3	49.5				ug/L	4	Standard
Br	81	8015.5	4.3				ug/L	1547	Standard
P	31	26808.7	3.1				ug/L	15286	Standard
S	34	2353.5	0.9				ug/L	1572	Standard
Sr	88	506.7	5.4				ug/L	73	Standard
C	12	183.3	20.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	3	Standard
Dy	164	4522.3	4.2				mg/L	6	Standard
Ho-1	165	2731.9	2.9				mg/L	10	Standard
Er	166	2333.5	8.8				mg/L	13	Standard
I	127	38263.4	7.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		126.735	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.114	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:55:02
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.544
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.921
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:55:02
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127901
Sample Date/Time: Wednesday, June 29, 2016 16:55:56
Number of Replicates: 3
Autosampler Position: 224
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94757.1	0.6				ug/L	74699	Standard
	Be	9	16.7	34.6	**0.0226**	0.005	23.2	ug/L	8	Standard
	Al	27	809583.1	0.4	**4.3294**	0.010	0.2	ug/L	1493	Standard
	Sc	45	70063.3	2.2				ug/L	61029	Standard
	Ti	47	239.7	2.8	**0.4418**	0.016	3.7	ug/L	35	Standard
	V	51	2635.2	3.3	**0.1083**	0.010	9.6	ug/L	1435	Standard
	Cr	52	10594.0	0.1	**0.2214**	0.004	1.7	ug/L	7635	Standard
	Cr	53	1061.7	10.6	**0.6708**	0.107	16.0	ug/L	363	Standard
	Mn	55	69843.0	1.9	**8.5894**	0.152	1.8	ug/L	747	Standard
	Co	59	472.7	1.3	**0.0251**	0.001	3.3	ug/L	234	Standard
	Ni	60	497.3	0.6	**0.1075**	0.002	1.4	ug/L	215	Standard
	Cu	65	314.7	2.4	**0.0436**	0.003	7.6	ug/L	104	Standard
	Zn	66	3346.7	3.2	**2.2142**	0.089	4.0	ug/L	363	Standard
>	**Ge**	72	568460.2	0.3				ug/L	495142	Standard
	As	75	-20.7	45.1	**0.0435**	0.007	16.6	ug/L	-50	Standard
	Se	82	18.3	17.6	**0.0072**	0.025	350.3	ug/L	14	Standard
	Se-1	77	58.0	6.2	**0.0694**	0.043	62.2	ug/L	42	Standard
>	**Ga**	71	123.3	16.9				mg/L	42	Standard
	Rb	85	875.0	6.6				ug/L	17	Standard
	Y	89	490331.7	1.2				ug/L	417098	Standard
>	**Rh**	103	16.7	45.8				ug/L	13	Standard
	Mo	98	175.2	3.1	**0.0513**	0.001	2.5	ug/L	11	Standard
	Ag	107	187.0	5.8	**0.0120**	0.001	10.3	ug/L	68	Standard
	Cd	111	8.4	36.2	**-0.0044**	0.001	27.6	mg/L	2	Standard
	Cd	114	20.0	63.9	**0.0013**	0.002	167.4	ug/L	12	Standard
>	**In**	115	676596.2	0.9				ug/L	587441	Standard
	Sn	118	621.7	13.0	**-0.0023**	0.012	549.8	ug/L	497	Standard
	Sb	123	148.0	17.5	**0.0233**	0.005	19.5	ug/L	78	Standard
	Ba	135	3597.4	2.3	**1.5099**	0.025	1.6	ug/L	24	Standard
	Ce	140	3133.7	5.7				ug/L	25	Standard
>	**Tb**	159	1053037.0	0.7				ug/L	939701	Standard
	Ho	165	73.3	37.6				ug/L	10	Standard
	Tl	203	38.7	9.8	**0.0024**	0.000	14.0	ug/L	4	Standard
	Tl	205	30.0	57.7	**0.0039**	0.002	55.8	ug/L	2	Standard
	Pb	206	500.3	0.4	**0.0113**	0.001	11.7	ug/L	240	Standard
	Pb	207	394.7	10.5	**0.0033**	0.008	256.3	ug/L	198	Standard
	Pb	208	1728.4	3.9	**0.0071**	0.003	37.0	ug/L	883	Standard
	U	238	362.3	7.5	**0.0420**	0.004	8.9	ug/L	3	Standard
>	**Bi**	209	594905.3	1.7				ug/L	523984	Standard

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:58:13
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.1214**	29.647	173.2	mg/L	0	Standard
Mg	24	156.7	18.1	**0.1107**	0.028	25.7	mg/L	37	Standard
K	39	30.0	60.1	**-0.0567**	0.034	59.1	mg/L	38	Standard
Ca	43	53.3	5.4	**0.4139**	0.213	51.5	mg/L	10	Standard
Fe	54	254.1	20.9	**0.0493**	0.038	78.1	mg/L	179	Standard
Fe	57	286.7	7.0	**-0.2249**	0.071	31.4	mg/L	335	Standard
> **Sc-1**	45	70063.3	2.2				mg/L	61029	Standard
Cl	35	33357.3	0.7				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	2703.6	11.1				ug/L	1547	Standard
P	31	21573.7	6.4				ug/L	15286	Standard
S	34	2518.5	1.8				ug/L	1572	Standard
Sr	88	90.0	20.0				ug/L	73	Standard
C	12	76.7	7.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	206.2	21.0				mg/L	6	Standard
Ho-1	165	73.3	37.6				mg/L	10	Standard
Er	166	80.0	12.5				mg/L	13	Standard
I	127	6318.0	20.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		126.852	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.807	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:58:13
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		115.177
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.535
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 16:59:09
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92097.4	0.2				ug/L	74699	Standard
	Be	9	59332.7	2.3	**54.6708**	1.340	2.5	ug/L	8	Standard
	Al	27	8920018.4	1.0	**49.2175**	0.589	1.2	ug/L	1493	Standard
	Sc	45	68374.0	2.7				ug/L	61029	Standard
	Ti	47	44746.6	0.7	**103.2626**	1.432	1.4	ug/L	35	Standard
	V	51	467593.6	0.9	**54.0107**	0.396	0.7	ug/L	1435	Standard
	Cr	52	444978.1	1.4	**51.5506**	1.111	2.2	ug/L	7635	Standard
	Cr	53	54491.2	2.0	**51.2973**	1.411	2.7	ug/L	363	Standard
	Mn	55	420710.6	2.8	**52.7641**	1.799	3.4	ug/L	747	Standard
	Co	59	425350.8	2.7	**56.3649**	1.831	3.2	ug/L	234	Standard
	Ni	60	109363.3	1.2	**51.0215**	0.971	1.9	ug/L	215	Standard
	Cu	65	108714.3	1.3	**51.0413**	1.018	2.0	ug/L	104	Standard
	Zn	66	63872.0	1.2	**50.7455**	1.020	2.0	ug/L	363	Standard
>	**Ge**	72	565538.7	0.9				ug/L	495142	Standard
	As	75	65590.6	1.0	**51.3298**	0.623	1.2	ug/L	-50	Standard
	Se	82	6676.5	1.4	**53.1245**	0.865	1.6	ug/L	14	Standard
	Se-1	77	4570.7	2.8	**54.1793**	1.989	3.7	ug/L	42	Standard
>	**Ga**	71	90.0	25.5				mg/L	42	Standard
	Rb	85	980.0	1.8				ug/L	17	Standard
	Y	89	481101.3	1.3				ug/L	417098	Standard
>	**Rh**	103	36.7	20.8				ug/L	13	Standard
	Mo	98	317636.1	2.3	**106.0178**	2.920	2.8	ug/L	11	Standard
	Ag	107	377887.3	1.0	**51.5689**	0.593	1.2	ug/L	68	Standard
	Cd	111	119799.3	1.9	**50.9674**	1.189	2.3	mg/L	2	Standard
	Cd	114	291345.5	1.8	**50.5384**	1.150	2.3	ug/L	12	Standard
>	**In**	115	660345.6	0.5				ug/L	587441	Standard
	Sn	118	333887.3	1.0	**50.7546**	0.756	1.5	ug/L	497	Standard
	Sb	123	263770.3	1.5	**50.2579**	0.987	2.0	ug/L	78	Standard
	Ba	135	110751.0	1.4	**48.1238**	0.926	1.9	ug/L	24	Standard
	Ce	140	100.0	13.2				ug/L	25	Standard
>	**Tb**	159	1027328.3	1.0				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	469667.3	0.3	**51.2282**	0.784	1.5	ug/L	4	Standard
	Tl	205	432915.4	1.7	**54.4135**	1.621	3.0	ug/L	2	Standard
	Pb	206	304012.7	1.3	**51.1045**	1.311	2.6	ug/L	240	Standard
	Pb	207	269196.3	1.3	**50.9328**	1.301	2.6	ug/L	198	Standard
	Pb	208	1105976.1	1.5	**52.9652**	1.506	2.8	ug/L	883	Standard
	U	238	408284.4	0.8	**53.3305**	1.093	2.0	ug/L	3	Standard
>	**Bi**	209	582837.4	1.3				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:01:26
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4664.1	3.2	**5.3139**	0.299	5.6	mg/L	37	Standard
K	39	2066.8	2.7	**3.9658**	0.176	4.4	mg/L	38	Standard
Ca	43	83.3	3.5	**2.1752**	0.111	5.1	mg/L	10	Standard
Fe	54	7259.4	3.7	**5.3547**	0.102	1.9	mg/L	179	Standard
Fe	57	2016.8	2.5	**5.0496**	0.304	6.0	mg/L	335	Standard
Sc-1	45	68374.0	2.7				mg/L	61029	Standard
Cl	35	31290.3	2.5				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	2203.5	6.4				ug/L	1547	Standard
P	31	22518.4	0.7				ug/L	15286	Standard
S	34	2795.3	6.4				ug/L	1572	Standard
Sr	88	58.3	13.1				ug/L	73	Standard
C	12	70.0	37.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	25.1	45.0				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	33.3	34.6				mg/L	13	Standard
I	127	4404.0	17.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	109.342		
Al	27	98.435		
Sc	45			
Ti	47	103.263		
V	51	108.021		
Cr	52	103.101		
Cr	53			
Mn	55	105.528		
Co	59	112.730		
Ni	60	102.043		
Cu	65	102.083		
Zn	66	101.491		
Ge	72		114.217	
As	75	102.660		
Se	82	106.249		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:01:26
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	106.018		
Ag	107	103.138		
Cd	111	101.935		
Cd	114			
In	115		112.411	
Sn	118	101.509		
Sb	123	100.516		
Ba	135	96.248		
Ce	140			
Tb	159			
Ho	165			
Tl	203	102.456		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.930		
U	238	106.661		
Bi	209		111.232	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:01:26
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 17:02:20
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92622.3	1.7				ug/L	74699	Standard
	Be	9	38.3	64.3	**0.0426**	0.023	52.8	ug/L	8	Standard
	Al	27	23370.4	151.2	**0.1159**	0.196	168.9	ug/L	1493	Standard
	Sc	45	68951.6	1.3				ug/L	61029	Standard
	Ti	47	40.3	21.1	**-0.0139**	0.019	138.6	ug/L	35	Standard
	V	51	1665.7	22.9	**0.0015**	0.044	2840.4	ug/L	1435	Standard
	Cr	52	7712.6	2.1	**-0.0923**	0.036	39.5	ug/L	7635	Standard
	Cr	53	583.3	10.9	**0.2339**	0.049	21.0	ug/L	363	Standard
	Mn	55	4624.3	138.3	**0.4433**	0.816	184.2	ug/L	747	Standard
	Co	59	398.3	65.9	**0.0165**	0.035	212.9	ug/L	234	Standard
	Ni	60	246.3	49.7	**-0.0062**	0.058	932.9	ug/L	215	Standard
	Cu	65	154.7	27.9	**-0.0294**	0.020	68.5	ug/L	104	Standard
	Zn	66	470.3	19.7	**-0.0693**	0.075	107.9	ug/L	363	Standard
>	**Ge**	72	553538.3	2.4				ug/L	495142	Standard
	As	75	-38.0	63.4	**0.0295**	0.019	62.8	ug/L	-50	Standard
	Se	82	20.6	12.7	**0.0307**	0.026	83.7	ug/L	14	Standard
	Se-1	77	63.0	12.0	**0.1490**	0.085	57.3	ug/L	42	Standard
>	**Ga**	71	71.7	14.5				mg/L	42	Standard
	Rb	85	25.0	0.0				ug/L	17	Standard
	Y	89	484828.2	0.4				ug/L	417098	Standard
>	**Rh**	103	11.7	65.5				ug/L	13	Standard
	Mo	98	326.7	26.1	**0.0989**	0.027	26.9	ug/L	11	Standard
	Ag	107	215.0	76.0	**0.0152**	0.021	139.6	ug/L	68	Standard
	Cd	111	16.1	49.4	**-0.0013**	0.003	242.1	mg/L	2	Standard
	Cd	114	45.0	78.6	**0.0054**	0.006	109.2	ug/L	12	Standard
>	**In**	115	687277.7	0.8				ug/L	587441	Standard
	Sn	118	931.7	18.6	**0.0418**	0.027	63.5	ug/L	497	Standard
	Sb	123	489.1	40.1	**0.0853**	0.036	41.8	ug/L	78	Standard
	Ba	135	52.0	47.0	**0.0057**	0.010	177.6	ug/L	24	Standard
	Ce	140	81.7	78.0				ug/L	25	Standard
>	**Tb**	159	1059717.9	1.6				ug/L	939701	Standard
	Ho	165	6.7	86.6				ug/L	10	Standard
	Tl	203	97.3	80.9	**0.0086**	0.008	97.3	ug/L	4	Standard
	Tl	205	126.7	77.5	**0.0158**	0.012	76.4	ug/L	2	Standard
	Pb	206	434.7	47.9	**0.0006**	0.034	5998.1	ug/L	240	Standard
	Pb	207	327.7	41.4	**-0.0091**	0.025	275.6	ug/L	198	Standard
	Pb	208	1259.4	26.8	**-0.0148**	0.016	107.4	ug/L	883	Standard
	U	238	94.7	104.8	**0.0077**	0.013	165.1	ug/L	3	Standard
>	**Bi**	209	593862.7	2.1				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:04:37
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	88.3	93.4	**0.0352**	0.093	262.9	mg/L	37	Standard
K	39	41.7	30.2	**-0.0325**	0.025	77.1	mg/L	38	Standard
Ca	43	21.7	58.1	**-1.3088**	0.708	54.1	mg/L	10	Standard
Fe	54	169.3	10.4	**-0.0113**	0.012	105.6	mg/L	179	Standard
Fe	57	255.0	7.1	**-0.3067**	0.061	20.0	mg/L	335	Standard
> **Sc-1**	45	68951.6	1.3				mg/L	61029	Standard
Cl	35	30862.1	1.0				ug/L	20195	Standard
Kr	83	3.7	31.5				ug/L	4	Standard
Br	81	2130.2	6.5				ug/L	1547	Standard
P	31	22533.4	1.3				ug/L	15286	Standard
S	34	2576.9	2.2				ug/L	1572	Standard
Sr	88	61.7	20.4				ug/L	73	Standard
C	12	70.0	62.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.1	95.9				mg/L	6	Standard
Ho-1	165	6.7	86.6				mg/L	10	Standard
Er	166	33.3	69.3				mg/L	13	Standard
I	127	3348.7	10.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.794	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:04:37
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.995
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.336
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Mn	55	

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:04:37
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127902 WG574001-03
Sample Date/Time: Wednesday, June 29, 2016 17:05:33
Number of Replicates: 3
Autosampler Position: 225
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	97844.9	1.2				ug/L	74699	Standard
	Be	9	628.3	5.4	**0.5525**	0.030	5.3	ug/L	8	Standard
	Al	27	67592650.7	0.8	**351.1766**	6.777	1.9	ug/L	1493	Standard
	Sc	45	76196.2	3.2				ug/L	61029	Standard
	Ti	47	35031.5	1.8	**85.4603**	1.454	1.7	ug/L	35	Standard
	V	51	127107.8	0.4	**15.3856**	0.087	0.6	ug/L	1435	Standard
	Cr	52	144594.0	0.9	**17.0395**	0.185	1.1	ug/L	7635	Standard
	Cr	53	17708.7	3.4	**17.4084**	0.576	3.3	ug/L	363	Standard
	Mn	55	9638149.6	0.5	**1281.4698**	6.113	0.5	ug/L	747	Standard
	Co	59	44221.4	1.1	**6.1626**	0.079	1.3	ug/L	234	Standard
	Ni	60	29440.2	0.7	**14.4336**	0.102	0.7	ug/L	215	Standard
	Cu	65	18715.2	1.1	**9.2058**	0.118	1.3	ug/L	104	Standard
	Zn	66	49911.7	1.0	**41.8483**	0.443	1.1	ug/L	363	Standard
>	**Ge**	72	534821.8	0.2				ug/L	495142	Standard
	As	75	5363.3	0.7	**4.4926**	0.039	0.9	ug/L	-50	Standard
	Se	82	81.7	4.2	**0.5509**	0.030	5.5	ug/L	14	Standard
	Se-1	77	90.0	5.1	**0.5184**	0.056	10.8	ug/L	42	Standard
>	**Ga**	71	10246.8	1.6				mg/L	42	Standard
	Rb	85	128176.6	1.2				ug/L	17	Standard
	Y	89	611826.2	1.2				ug/L	417098	Standard
>	**Rh**	103	203.3	5.1				ug/L	13	Standard
	Mo	98	8459.2	1.1	**2.8837**	0.063	2.2	ug/L	11	Standard
	Ag	107	335.7	32.7	**0.0339**	0.015	43.7	ug/L	68	Standard
	Cd	111	167.4	13.2	**0.0649**	0.009	13.6	mg/L	2	Standard
	Cd	114	471.4	29.4	**0.0814**	0.024	29.1	ug/L	12	Standard
>	**In**	115	645352.1	1.1				ug/L	587441	Standard
	Sn	118	2433.5	1.0	**0.2845**	0.006	2.0	ug/L	497	Standard
	Sb	123	747.2	5.1	**0.1415**	0.006	4.4	ug/L	78	Standard
	Ba	135	186708.8	1.1	**83.0208**	0.490	0.6	ug/L	24	Standard
	Ce	140	555105.9	0.4				ug/L	25	Standard
>	**Tb**	159	1017797.9	1.6				ug/L	939701	Standard
	Ho	165	10682.1	1.8				ug/L	10	Standard
	Tl	203	801.4	12.7	**0.0906**	0.011	12.4	ug/L	4	Standard
	Tl	205	726.7	5.1	**0.0967**	0.005	5.6	ug/L	2	Standard
	Pb	206	28687.5	0.5	**5.0315**	0.023	0.5	ug/L	240	Standard
	Pb	207	23046.8	1.0	**4.5438**	0.005	0.1	ug/L	198	Standard
	Pb	208	100519.8	0.6	**5.0199**	0.074	1.5	ug/L	883	Standard
	U	238	32058.9	0.6	**4.4211**	0.050	1.1	ug/L	3	Standard
>	**Bi**	209	551473.5	0.9				ug/L	523984	Standard

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:07:50
Page 1

Approved: July 04, 2016

Na	23	16.7	17.3	**153.4719**	22.118	14.4	mg/L	0	Standard
Mg	24	8050.5	4.4	**8.2571**	0.176	2.1	mg/L	37	Standard
K	39	990.0	1.3	**1.6393**	0.041	2.5	mg/L	38	Standard
Ca	43	1383.4	8.1	**67.3924**	3.470	5.1	mg/L	10	Standard
Fe	54	12019.5	0.9	**8.0301**	0.303	3.8	mg/L	179	Standard
Fe	57	3888.8	1.5	**9.5221**	0.470	4.9	mg/L	335	Standard
> **Sc-1**	45	76196.2	3.2				mg/L	61029	Standard
Cl	35	38282.3	2.8				ug/L	20195	Standard
Kr	83	1.7	124.9				ug/L	4	Standard
Br	81	9239.5	4.1				ug/L	1547	Standard
P	31	33428.5	3.8				ug/L	15286	Standard
S	34	2046.8	10.4				ug/L	1572	Standard
Sr	88	885.0	8.4				ug/L	73	Standard
C	12	223.3	22.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	3	Standard
Dy	164	18454.9	2.8				mg/L	6	Standard
Ho-1	165	10682.1	1.8				mg/L	10	Standard
Er	166	9956.6	4.7				mg/L	13	Standard
I	127	65381.5	9.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		130.986	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.014	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:07:50
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	109.858
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	105.246
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:07:50
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127902 WG574001-03
Sample Date/Time: Wednesday, June 29, 2016 17:08:45
Number of Replicates: 3
Autosampler Position: 226
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	97333.3	1.1				ug/L	74699	Standard
	Be	9	23.3	32.7	**0.0279**	0.007	23.3	ug/L	8	Standard
	Al	27	1471688.9	2.0	**7.6717**	0.099	1.3	ug/L	1493	Standard
	Sc	45	68584.9	0.8				ug/L	61029	Standard
	Ti	47	736.0	3.1	**1.5714**	0.076	4.9	ug/L	35	Standard
	V	51	4360.2	2.0	**0.3040**	0.018	5.9	ug/L	1435	Standard
	Cr	52	10891.2	1.4	**0.2475**	0.011	4.5	ug/L	7635	Standard
	Cr	53	901.7	3.1	**0.5139**	0.018	3.4	ug/L	363	Standard
	Mn	55	209342.7	1.0	**25.8667**	0.657	2.5	ug/L	747	Standard
	Co	59	1184.0	8.5	**0.1180**	0.016	13.3	ug/L	234	Standard
	Ni	60	823.4	2.4	**0.2566**	0.015	5.7	ug/L	215	Standard
	Cu	65	604.0	5.8	**0.1771**	0.017	9.7	ug/L	104	Standard
	Zn	66	3536.4	2.1	**2.3459**	0.038	1.6	ug/L	363	Standard
>	**Ge**	72	572433.6	1.6				ug/L	495142	Standard
	As	75	56.5	19.8	**0.1032**	0.008	8.2	ug/L	-50	Standard
	Se	82	22.7	8.9	**0.0409**	0.015	36.5	ug/L	14	Standard
	Se-1	77	56.3	3.7	**0.0451**	0.032	71.6	ug/L	42	Standard
>	**Ga**	71	286.7	20.8				mg/L	42	Standard
	Rb	85	2816.9	8.1				ug/L	17	Standard
	Y	89	483993.7	0.1				ug/L	417098	Standard
>	**Rh**	103	11.7	65.5				ug/L	13	Standard
	Mo	98	255.8	9.6	**0.0784**	0.008	10.6	ug/L	11	Standard
	Ag	107	215.7	39.7	**0.0161**	0.012	72.2	ug/L	68	Standard
	Cd	111	33.2	125.7	**0.0061**	0.018	290.4	mg/L	2	Standard
	Cd	114	83.8	140.0	**0.0123**	0.020	164.2	ug/L	12	Standard
>	**In**	115	670122.5	0.4				ug/L	587441	Standard
	Sn	118	730.0	9.6	**0.0149**	0.011	73.4	ug/L	497	Standard
	Sb	123	198.5	22.7	**0.0331**	0.009	26.1	ug/L	78	Standard
	Ba	135	4104.6	4.5	**1.7422**	0.086	5.0	ug/L	24	Standard
	Ce	140	11919.7	3.4				ug/L	25	Standard
>	**Tb**	159	1040073.5	0.3				ug/L	939701	Standard
	Ho	165	248.3	25.2				ug/L	10	Standard
	Tl	203	119.3	122.5	**0.0111**	0.016	142.6	ug/L	4	Standard
	Tl	205	45.0	44.4	**0.0058**	0.003	43.8	ug/L	2	Standard
	Pb	206	1027.4	11.1	**0.0985**	0.020	20.7	ug/L	240	Standard
	Pb	207	828.0	1.8	**0.0839**	0.004	4.9	ug/L	198	Standard
	Pb	208	3389.1	2.4	**0.0853**	0.005	6.3	ug/L	883	Standard
	U	238	690.0	3.8	**0.0840**	0.004	4.3	ug/L	3	Standard
>	**Bi**	209	594228.5	1.0				ug/L	523984	Standard

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:11:02
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	215.0	6.2	**0.1819**	0.016	8.8	mg/L	37	Standard
K	39	53.3	14.3	**-0.0093**	0.014	153.2	mg/L	38	Standard
Ca	43	48.3	15.8	**0.1914**	0.411	214.9	mg/L	10	Standard
Fe	54	443.5	15.8	**0.1966**	0.055	28.1	mg/L	179	Standard
Fe	57	323.3	11.7	**-0.0960**	0.118	122.9	mg/L	335	Standard
> **Sc-1**	45	68584.9	0.8				mg/L	61029	Standard
Cl	35	29711.8	0.5				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	2166.8	7.4				ug/L	1547	Standard
P	31	19419.1	3.8				ug/L	15286	Standard
S	34	2206.8	3.0				ug/L	1572	Standard
Sr	88	96.7	19.6				ug/L	73	Standard
C	12	46.7	49.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	390.3	4.3				mg/L	6	Standard
Ho-1	165	248.3	25.2				mg/L	10	Standard
Er	166	203.3	36.3				mg/L	13	Standard
I	127	5617.8	25.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		130.301	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.610	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:11:02
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.075
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.406
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127903S WG574001-09
Sample Date/Time: Wednesday, June 29, 2016 17:11:56
Number of Replicates: 3
Autosampler Position: 227
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	103734.0	1.8				ug/L	74699	Standard
	Be	9	62659.4	2.5	**51.2580**	0.858	1.7	ug/L	8	Standard
	Al	27	65559814.6	2.9	**321.3254**	12.360	3.8	ug/L	1493	Standard
	Sc	45	78765.9	1.8				ug/L	61029	Standard
	Ti	47	33652.0	1.4	**81.0287**	2.211	2.7	ug/L	35	Standard
	V	51	549023.7	0.9	**66.2300**	1.368	2.1	ug/L	1435	Standard
	Cr	52	510507.6	0.8	**61.9303**	1.817	2.9	ug/L	7635	Standard
	Cr	53	62490.3	1.3	**61.4602**	1.770	2.9	ug/L	363	Standard
	Mn	55	9311676.0	0.4	**1221.9970**	25.998	2.1	ug/L	747	Standard
	Co	59	420012.9	0.8	**58.0870**	1.492	2.6	ug/L	234	Standard
	Ni	60	121605.8	1.9	**59.2404**	2.250	3.8	ug/L	215	Standard
	Cu	65	110409.9	1.5	**54.1128**	1.822	3.4	ug/L	104	Standard
	Zn	66	101498.2	0.9	**84.4641**	2.491	2.9	ug/L	363	Standard
>	**Ge**	72	542031.1	2.4				ug/L	495142	Standard
	As	75	66455.1	1.4	**54.2648**	0.548	1.0	ug/L	-50	Standard
	Se	82	6239.8	1.8	**51.8120**	1.305	2.5	ug/L	14	Standard
	Se-1	77	4136.2	2.8	**51.1231**	1.591	3.1	ug/L	42	Standard
>	**Ga**	71	10863.9	0.3				mg/L	42	Standard
	Rb	85	119392.1	2.4				ug/L	17	Standard
	Y	89	586391.1	1.2				ug/L	417098	Standard
>	**Rh**	103	175.0	15.1				ug/L	13	Standard
	Mo	98	5740.7	2.3	**1.9719**	0.058	3.0	ug/L	11	Standard
	Ag	107	333249.6	2.0	**46.9378**	1.203	2.6	ug/L	68	Standard
	Cd	111	109090.4	0.8	**47.8987**	0.737	1.5	mg/L	2	Standard
	Cd	114	263561.8	0.5	**47.1881**	0.999	2.1	ug/L	12	Standard
>	**In**	115	639905.2	1.9				ug/L	587441	Standard
	Sn	118	1925.1	5.8	**0.2078**	0.017	8.3	ug/L	497	Standard
	Sb	123	124661.9	1.5	**24.5159**	0.713	2.9	ug/L	78	Standard
	Ba	135	268706.4	1.4	**120.5245**	2.161	1.8	ug/L	24	Standard
	Ce	140	538323.1	0.7				ug/L	25	Standard
>	**Tb**	159	1045169.0	2.0				ug/L	939701	Standard
	Ho	165	11317.6	5.4				ug/L	10	Standard
	Tl	203	438870.4	1.4	**49.1992**	0.166	0.3	ug/L	4	Standard
	Tl	205	398898.5	1.5	**51.5243**	0.281	0.5	ug/L	2	Standard
	Pb	206	314887.1	0.5	**54.4073**	0.494	0.9	ug/L	240	Standard
	Pb	207	267442.5	0.8	**52.0088**	0.664	1.3	ug/L	198	Standard
	Pb	208	1101699.4	1.8	**54.2208**	0.536	1.0	ug/L	883	Standard
	U	238	415059.0	1.7	**55.7208**	0.743	1.3	ug/L	3	Standard
>	**Bi**	209	566992.1	1.1				ug/L	523984	Standard

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:14:13
Page 1

Approved: July 04, 2016

Na	23	8.3	69.3	**74.8950**	52.957	70.7	mg/L	0	Standard
Mg	24	7490.2	5.9	**7.4319**	0.549	7.4	mg/L	37	Standard
K	39	995.0	6.5	**1.5912**	0.134	8.4	mg/L	38	Standard
Ca	43	1293.4	1.8	**60.7979**	1.957	3.2	mg/L	10	Standard
Fe	54	10550.1	1.1	**6.7940**	0.164	2.4	mg/L	179	Standard
Fe	57	3448.7	1.2	**8.0106**	0.227	2.8	mg/L	335	Standard
> **Sc-1**	45	78765.9	1.8				mg/L	61029	Standard
Cl	35	36168.5	1.3				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	7995.5	8.9				ug/L	1547	Standard
P	31	34405.7	2.4				ug/L	15286	Standard
S	34	1880.1	7.8				ug/L	1572	Standard
Sr	88	881.7	7.6				ug/L	73	Standard
C	12	140.0	46.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	3	Standard
Dy	164	18849.9	4.6				mg/L	6	Standard
Ho-1	165	11317.6	5.4				mg/L	10	Standard
Er	166	9863.2	2.6				mg/L	13	Standard
I	127	61177.5	10.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		138.870	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		109.470	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:14:13
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.931
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.208
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:14:13
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127903S WG574001-09
Sample Date/Time: Wednesday, June 29, 2016 17:15:08
Number of Replicates: 3
Autosampler Position: 228
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100014.2	2.8				ug/L	74699	Standard
	Be	9	1316.7	11.4	**1.1231**	0.100	8.9	ug/L	8	Standard
	Al	27	1420391.6	2.6	**7.2057**	0.041	0.6	ug/L	1493	Standard
	Sc	45	65355.8	3.5				ug/L	61029	Standard
	Ti	47	751.7	2.2	**1.7435**	0.043	2.4	ug/L	35	Standard
	V	51	12553.2	2.5	**1.3568**	0.030	2.2	ug/L	1435	Standard
	Cr	52	18458.6	1.1	**1.3028**	0.035	2.7	ug/L	7635	Standard
	Cr	53	1771.8	3.3	**1.4619**	0.098	6.7	ug/L	363	Standard
	Mn	55	203580.1	1.5	**27.1635**	0.320	1.2	ug/L	747	Standard
	Co	59	9135.1	1.0	**1.2550**	0.021	1.7	ug/L	234	Standard
	Ni	60	2779.3	1.4	**1.2628**	0.031	2.5	ug/L	215	Standard
	Cu	65	2563.2	1.7	**1.1828**	0.027	2.3	ug/L	104	Standard
	Zn	66	4447.0	0.1	**3.3488**	0.082	2.5	ug/L	363	Standard
>	**Ge**	72	530183.0	2.3				ug/L	495142	Standard
	As	75	1256.2	4.8	**1.1077**	0.065	5.9	ug/L	-50	Standard
	Se	82	144.5	11.0	**1.0936**	0.160	14.6	ug/L	14	Standard
	Se-1	77	149.3	5.6	**1.2873**	0.101	7.8	ug/L	42	Standard
>	**Ga**	71	291.7	12.0				mg/L	42	Standard
	Rb	85	2473.5	8.7				ug/L	17	Standard
	Y	89	480063.9	1.1				ug/L	417098	Standard
>	**Rh**	103	18.3	15.7				ug/L	13	Standard
	Mo	98	168.4	5.0	**0.0496**	0.003	5.9	ug/L	11	Standard
	Ag	107	7406.8	1.4	**0.9815**	0.012	1.2	ug/L	68	Standard
	Cd	111	2324.9	2.0	**0.9649**	0.013	1.3	mg/L	2	Standard
	Cd	114	5499.5	2.5	**0.9362**	0.028	2.9	ug/L	12	Standard
>	**In**	115	671407.0	1.0				ug/L	587441	Standard
	Sn	118	800.0	6.0	**0.0251**	0.007	29.1	ug/L	497	Standard
	Sb	123	3310.0	4.5	**0.6160**	0.025	4.0	ug/L	78	Standard
	Ba	135	5738.4	0.8	**2.4371**	0.009	0.4	ug/L	24	Standard
	Ce	140	11352.6	1.4				ug/L	25	Standard
>	**Tb**	159	1030086.3	0.3				ug/L	939701	Standard
	Ho	165	295.0	7.8				ug/L	10	Standard
	Tl	203	9225.5	0.2	**0.9995**	0.013	1.3	ug/L	4	Standard
	Tl	205	8459.0	5.0	**1.0575**	0.041	3.9	ug/L	2	Standard
	Pb	206	6958.9	0.8	**1.0943**	0.025	2.3	ug/L	240	Standard
	Pb	207	5899.8	0.2	**1.0421**	0.014	1.4	ug/L	198	Standard
	Pb	208	25022.5	1.7	**1.1197**	0.025	2.3	ug/L	883	Standard
	U	238	8748.9	2.2	**1.1326**	0.025	2.2	ug/L	3	Standard
>	**Bi**	209	585775.3	1.4				ug/L	523984	Standard

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:17:25
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	185.0	5.4	**0.1579**	0.009	5.9	mg/L	37	Standard
K	39	46.7	32.7	**-0.0170**	0.035	202.9	mg/L	38	Standard
Ca	43	53.3	5.4	**0.6286**	0.271	43.1	mg/L	10	Standard
Fe	54	400.2	11.9	**0.1796**	0.048	26.7	mg/L	179	Standard
Fe	57	335.0	9.8	**-0.0119**	0.080	674.7	mg/L	335	Standard
Sc-1	45	65355.8	3.5				mg/L	61029	Standard
Cl	35	29320.3	1.0				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	2030.1	9.5				ug/L	1547	Standard
P	31	19806.3	2.6				ug/L	15286	Standard
S	34	2145.2	0.2				ug/L	1572	Standard
Sr	88	86.7	23.3				ug/L	73	Standard
C	12	80.0	66.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	379.6	21.3				mg/L	6	Standard
Ho-1	165	295.0	7.8				mg/L	10	Standard
Er	166	216.7	11.6				mg/L	13	Standard
I	127	5157.6	28.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		133.890	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		107.077	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:17:25
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.294
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.792
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127904SD WG574001-10
Sample Date/Time: Wednesday, June 29, 2016 17:18:19
Number of Replicates: 3
Autosampler Position: 229
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100168.4	0.8				ug/L	74699	Standard
	Be	9	61913.0	2.3	**52.4502**	1.133	2.2	ug/L	8	Standard
	Al	27	66193036.9	1.4	**335.9039**	6.576	2.0	ug/L	1493	Standard
	Sc	45	75343.6	2.4				ug/L	61029	Standard
	Ti	47	40945.9	1.4	**103.9053**	1.926	1.9	ug/L	35	Standard
	V	51	552663.1	1.5	**70.2550**	1.056	1.5	ug/L	1435	Standard
	Cr	52	495290.7	1.7	**63.3235**	1.419	2.2	ug/L	7635	Standard
	Cr	53	60987.5	1.0	**63.2038**	0.948	1.5	ug/L	363	Standard
	Mn	55	9453870.8	1.1	**1307.2105**	20.907	1.6	ug/L	747	Standard
	Co	59	421300.3	0.1	**61.3867**	0.363	0.6	ug/L	234	Standard
	Ni	60	119269.8	2.5	**61.2106**	1.745	2.9	ug/L	215	Standard
	Cu	65	106997.2	1.7	**55.2474**	1.210	2.2	ug/L	104	Standard
	Zn	66	101243.5	1.4	**88.7847**	1.659	1.9	ug/L	363	Standard
>	**Ge**	72	514295.3	0.5				ug/L	495142	Standard
	As	75	64887.3	0.7	**55.8337**	0.652	1.2	ug/L	-50	Standard
	Se	82	6128.3	0.4	**53.6210**	0.432	0.8	ug/L	14	Standard
	Se-1	77	4025.2	1.7	**52.4403**	1.142	2.2	ug/L	42	Standard
>	**Ga**	71	13474.3	0.3				mg/L	42	Standard
	Rb	85	148814.4	3.2				ug/L	17	Standard
	Y	89	590511.6	1.2				ug/L	417098	Standard
>	**Rh**	103	170.0	19.3				ug/L	13	Standard
	Mo	98	5249.8	1.8	**1.8847**	0.047	2.5	ug/L	11	Standard
	Ag	107	319729.4	0.9	**47.0680**	0.306	0.7	ug/L	68	Standard
	Cd	111	105379.0	1.4	**48.3587**	0.323	0.7	mg/L	2	Standard
	Cd	114	254720.6	1.7	**47.6605**	0.543	1.1	ug/L	12	Standard
>	**In**	115	612131.8	1.0				ug/L	587441	Standard
	Sn	118	2201.8	4.9	**0.2670**	0.017	6.2	ug/L	497	Standard
	Sb	123	135513.9	0.7	**27.8528**	0.396	1.4	ug/L	78	Standard
	Ba	135	276700.5	0.9	**129.7213**	0.066	0.1	ug/L	24	Standard
	Ce	140	594096.9	1.6				ug/L	25	Standard
>	**Tb**	159	982831.9	0.9				ug/L	939701	Standard
	Ho	165	12233.3	1.6				ug/L	10	Standard
	Tl	203	424338.5	0.5	**50.3728**	0.553	1.1	ug/L	4	Standard
	Tl	205	393687.9	2.5	**53.8449**	1.302	2.4	ug/L	2	Standard
	Pb	206	310823.5	0.5	**56.8661**	0.171	0.3	ug/L	240	Standard
	Pb	207	265328.4	0.5	**54.6347**	0.327	0.6	ug/L	198	Standard
	Pb	208	1090751.7	0.6	**56.8486**	0.519	0.9	ug/L	883	Standard
	U	238	409441.4	1.4	**58.2033**	0.887	1.5	ug/L	3	Standard
>	**Bi**	209	535481.7	0.7				ug/L	523984	Standard

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:20:36
Page 1

Approved: July 04, 2016

Na	23	6.7	86.6	**61.5240**	53.288	86.6	mg/L	0	Standard
Mg	24	7358.5	1.6	**7.6312**	0.188	2.5	mg/L	37	Standard
K	39	1028.4	3.5	**1.7267**	0.031	1.8	mg/L	38	Standard
Ca	43	1340.1	2.9	**66.0398**	1.115	1.7	mg/L	10	Standard
Fe	54	11628.0	2.4	**7.8474**	0.058	0.7	mg/L	179	Standard
Fe	57	3898.8	5.2	**9.6559**	0.353	3.7	mg/L	335	Standard
Sc-1	45	75343.6	2.4				mg/L	61029	Standard
Cl	35	34624.2	2.4				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	10260.1	1.8				ug/L	1547	Standard
P	31	37270.8	0.5				ug/L	15286	Standard
S	34	1845.1	3.3				ug/L	1572	Standard
Sr	88	923.4	9.7				ug/L	73	Standard
C	12	203.3	29.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	16.7	91.7				mg/L	3	Standard
Dy	164	20358.1	1.7				mg/L	6	Standard
Ho-1	165	12233.3	1.6				mg/L	10	Standard
Er	166	11334.2	2.0				mg/L	13	Standard
I	127	62766.8	7.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		134.096	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		103.868	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:20:36
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	104.203
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.194
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Ti 47 Upper, S, EEE	Ti	47	

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:20:36
Page 3

Approved: July 04, 2016

Mn 55 Upper, S, EEE	Mn	55
Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127904SD WG574001-10
Sample Date/Time: Wednesday, June 29, 2016 17:21:31
Number of Replicates: 3
Autosampler Position: 230
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96610.6	3.3				ug/L	74699	Standard
	Be	9	1288.4	4.7	**1.1394**	0.049	4.3	ug/L	8	Standard
	Al	27	1461113.0	0.9	**7.6780**	0.187	2.4	ug/L	1493	Standard
	Sc	45	62711.5	6.5				ug/L	61029	Standard
	Ti	47	920.7	1.4	**2.1467**	0.013	0.6	ug/L	35	Standard
	V	51	13378.9	3.6	**1.4502**	0.085	5.9	ug/L	1435	Standard
	Cr	52	18691.9	1.7	**1.3185**	0.044	3.3	ug/L	7635	Standard
	Cr	53	1716.8	3.7	**1.3959**	0.091	6.5	ug/L	363	Standard
	Mn	55	207917.4	1.7	**27.5865**	0.619	2.2	ug/L	747	Standard
	Co	59	9044.0	2.3	**1.2346**	0.031	2.5	ug/L	234	Standard
	Ni	60	2824.3	2.7	**1.2771**	0.044	3.5	ug/L	215	Standard
	Cu	65	2601.9	3.8	**1.1945**	0.050	4.2	ug/L	104	Standard
	Zn	66	4831.1	2.2	**3.6537**	0.142	3.9	ug/L	363	Standard
>	**Ge**	72	533228.6	1.6				ug/L	495142	Standard
	As	75	1337.8	3.9	**1.1692**	0.061	5.2	ug/L	-50	Standard
	Se	82	140.1	2.8	**1.0475**	0.027	2.6	ug/L	14	Standard
	Se-1	77	133.7	3.0	**1.0773**	0.056	5.2	ug/L	42	Standard
>	**Ga**	71	393.3	11.4				mg/L	42	Standard
	Rb	85	3227.0	1.4				ug/L	17	Standard
	Y	89	450037.4	1.9				ug/L	417098	Standard
>	**Rh**	103	18.3	87.7				ug/L	13	Standard
	Mo	98	147.9	13.7	**0.0463**	0.009	18.8	ug/L	11	Standard
	Ag	107	7379.1	1.6	**1.0451**	0.031	3.0	ug/L	68	Standard
	Cd	111	2296.2	2.5	**1.0190**	0.058	5.7	mg/L	2	Standard
	Cd	114	5492.1	3.4	**0.9981**	0.024	2.4	ug/L	12	Standard
>	**In**	115	629018.5	3.2				ug/L	587441	Standard
	Sn	118	580.0	12.2	**-0.0020**	0.010	499.8	ug/L	497	Standard
	Sb	123	3693.7	7.1	**0.7349**	0.053	7.2	ug/L	78	Standard
	Ba	135	5980.2	1.6	**2.7136**	0.046	1.7	ug/L	24	Standard
	Ce	140	12665.3	1.3				ug/L	25	Standard
>	**Tb**	159	1007879.7	3.3				ug/L	939701	Standard
	Ho	165	265.0	7.5				ug/L	10	Standard
	Tl	203	9323.2	2.1	**1.0256**	0.019	1.8	ug/L	4	Standard
	Tl	205	8567.4	1.0	**1.0883**	0.033	3.0	ug/L	2	Standard
	Pb	206	7313.4	2.1	**1.1727**	0.042	3.6	ug/L	240	Standard
	Pb	207	6050.9	2.1	**1.0885**	0.043	4.0	ug/L	198	Standard
	Pb	208	24824.3	2.0	**1.1284**	0.015	1.4	ug/L	883	Standard
	U	238	8436.0	2.0	**1.1088**	0.013	1.2	ug/L	3	Standard
>	**Bi**	209	577014.5	3.2				ug/L	523984	Standard

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:23:48
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	200.0	10.9	**0.1869**	0.033	17.5	mg/L	37	Standard
K	39	61.7	46.1	**0.0210**	0.066	315.1	mg/L	38	Standard
Ca	43	46.7	12.4	**0.3502**	0.375	107.1	mg/L	10	Standard
Fe	54	378.5	13.7	**0.1751**	0.049	28.1	mg/L	179	Standard
Fe	57	365.0	16.4	**0.1419**	0.250	175.9	mg/L	335	Standard
> **Sc-1**	45	62711.5	6.5				mg/L	61029	Standard
Cl	35	27760.1	3.1				ug/L	20195	Standard
Kr	83	3.3	45.8				ug/L	4	Standard
Br	81	2080.1	7.5				ug/L	1547	Standard
P	31	18322.7	2.7				ug/L	15286	Standard
S	34	1913.5	2.4				ug/L	1572	Standard
Sr	88	98.3	10.6				ug/L	73	Standard
C	12	103.3	31.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	615.8	1.7				mg/L	6	Standard
Ho-1	165	265.0	7.5				mg/L	10	Standard
Er	166	296.7	10.3				mg/L	13	Standard
I	127	5109.3	32.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		129.334	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		107.692	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:23:48
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		107.078
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.121
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127905
Sample Date/Time: Wednesday, June 29, 2016 17:24:42
Number of Replicates: 3
Autosampler Position: 231
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98923.6	1.9				ug/L	74699	Standard
	Be	9	425.0	5.1	**0.3722**	0.020	5.4	ug/L	8	Standard
	Al	27	40729181.9	0.9	**209.2967**	3.040	1.5	ug/L	1493	Standard
	Sc	45	74965.2	4.5				ug/L	61029	Standard
	Ti	47	31002.7	2.7	**78.7018**	1.731	2.2	ug/L	35	Standard
	V	51	100570.7	0.4	**12.6354**	0.094	0.7	ug/L	1435	Standard
	Cr	52	71317.5	0.6	**8.2486**	0.021	0.3	ug/L	7635	Standard
	Cr	53	8882.6	3.6	**8.9299**	0.295	3.3	ug/L	363	Standard
	Mn	55	1183907.3	1.1	**163.7014**	1.986	1.2	ug/L	747	Standard
	Co	59	16826.7	0.9	**2.4180**	0.026	1.1	ug/L	234	Standard
	Ni	60	14706.8	1.9	**7.4444**	0.110	1.5	ug/L	215	Standard
	Cu	65	7822.7	1.1	**3.9461**	0.031	0.8	ug/L	104	Standard
	Zn	66	28811.4	1.3	**24.9594**	0.201	0.8	ug/L	363	Standard
>	**Ge**	72	513872.0	0.5				ug/L	495142	Standard
	As	75	2302.8	2.2	**2.0406**	0.054	2.6	ug/L	-50	Standard
	Se	82	647.6	3.3	**5.5477**	0.199	3.6	ug/L	14	Standard
	Se-1	77	492.7	1.9	**5.8772**	0.106	1.8	ug/L	42	Standard
>	**Ga**	71	9201.1	2.2				mg/L	42	Standard
	Rb	85	118130.4	0.8				ug/L	17	Standard
	Y	89	544220.1	0.8				ug/L	417098	Standard
>	**Rh**	103	83.3	19.3				ug/L	13	Standard
	Mo	98	4972.1	0.8	**1.6995**	0.011	0.6	ug/L	11	Standard
	Ag	107	221.0	6.7	**0.0181**	0.002	11.5	ug/L	68	Standard
	Cd	111	99.4	11.4	**0.0355**	0.005	13.9	mg/L	2	Standard
	Cd	114	258.8	15.4	**0.0440**	0.007	16.4	ug/L	12	Standard
>	**In**	115	642627.7	0.7				ug/L	587441	Standard
	Sn	118	4954.1	4.3	**0.6806**	0.035	5.1	ug/L	497	Standard
	Sb	123	697.4	10.3	**0.1324**	0.015	11.0	ug/L	78	Standard
	Ba	135	175907.0	0.9	**78.5481**	0.493	0.6	ug/L	24	Standard
	Ce	140	308696.3	0.2				ug/L	25	Standard
>	**Tb**	159	1018138.8	1.9				ug/L	939701	Standard
	Ho	165	5902.8	0.8				ug/L	10	Standard
	Tl	203	694.0	1.8	**0.0782**	0.000	0.4	ug/L	4	Standard
	Tl	205	615.0	15.7	**0.0817**	0.011	14.1	ug/L	2	Standard
	Pb	206	15714.8	1.1	**2.7233**	0.017	0.6	ug/L	240	Standard
	Pb	207	12858.1	1.8	**2.5034**	0.026	1.0	ug/L	198	Standard
	Pb	208	54664.4	0.6	**2.6954**	0.031	1.2	ug/L	883	Standard
	U	238	8998.7	0.4	**1.2376**	0.021	1.7	ug/L	3	Standard
>	**Bi**	209	551643.9	1.7				ug/L	523984	Standard

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:26:59
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**46.3980**	45.060	97.1	mg/L	0	Standard
Mg	24	7338.4	2.5	**7.6542**	0.303	4.0	mg/L	37	Standard
K	39	668.3	6.0	**1.0889**	0.064	5.9	mg/L	38	Standard
Ca	43	676.7	13.4	**32.4506**	6.121	18.9	mg/L	10	Standard
Fe	54	6512.7	2.4	**4.3601**	0.118	2.7	mg/L	179	Standard
Fe	57	2150.2	3.4	**4.8854**	0.383	7.8	mg/L	335	Standard
> **Sc-1**	45	74965.2	4.5				mg/L	61029	Standard
Cl	35	28544.2	1.7				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	11494.4	5.7				ug/L	1547	Standard
P	31	34679.6	1.6				ug/L	15286	Standard
S	34	2073.5	2.2				ug/L	1572	Standard
Sr	88	356.7	8.9				ug/L	73	Standard
C	12	163.3	55.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	3	Standard
Dy	164	9835.4	4.4				mg/L	6	Standard
Ho-1	165	5902.8	0.8				mg/L	10	Standard
Er	166	5037.5	9.2				mg/L	13	Standard
I	127	62917.2	4.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.430	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.783	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:26:59
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.394
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.279
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:26:59
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127905
Sample Date/Time: Wednesday, June 29, 2016 17:27:54
Number of Replicates: 3
Autosampler Position: 232
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95902.7	2.3				ug/L	74699	Standard
	Be	9	26.7	54.1	**0.0312**	0.013	41.0	ug/L	8	Standard
	Al	27	866639.8	0.6	**4.5819**	0.130	2.8	ug/L	1493	Standard
	Sc	45	64717.9	1.3				ug/L	61029	Standard
	Ti	47	665.3	5.2	**1.4927**	0.099	6.6	ug/L	35	Standard
	V	51	3898.6	6.9	**0.2752**	0.030	11.1	ug/L	1435	Standard
	Cr	52	9682.4	1.6	**0.1679**	0.023	13.9	ug/L	7635	Standard
	Cr	53	670.0	10.3	**0.3316**	0.072	21.7	ug/L	363	Standard
	Mn	55	29584.3	8.9	**3.7268**	0.331	8.9	ug/L	747	Standard
	Co	59	652.3	12.4	**0.0528**	0.011	20.6	ug/L	234	Standard
	Ni	60	547.0	3.7	**0.1426**	0.010	7.1	ug/L	215	Standard
	Cu	65	341.3	5.6	**0.0636**	0.011	17.0	ug/L	104	Standard
	Zn	66	2764.3	3.0	**1.8532**	0.047	2.5	ug/L	363	Standard
>	**Ge**	72	542890.6	1.0				ug/L	495142	Standard
	As	75	11.1	320.4	**0.0688**	0.029	42.5	ug/L	-50	Standard
	Se	82	31.5	11.6	**0.1237**	0.032	26.0	ug/L	14	Standard
	Se-1	77	57.7	24.7	**0.0971**	0.173	178.1	ug/L	42	Standard
>	**Ga**	71	253.3	16.8				mg/L	42	Standard
	Rb	85	2388.5	2.4				ug/L	17	Standard
	Y	89	480185.0	1.7				ug/L	417098	Standard
>	**Rh**	103	16.7	34.6				ug/L	13	Standard
	Mo	98	144.5	13.4	**0.0432**	0.007	17.2	ug/L	11	Standard
	Ag	107	273.0	44.6	**0.0250**	0.018	70.2	ug/L	68	Standard
	Cd	111	17.1	70.4	**-0.0005**	0.005	1058.7	mg/L	2	Standard
	Cd	114	57.8	101.1	**0.0081**	0.010	129.1	ug/L	12	Standard
>	**In**	115	652419.9	1.9				ug/L	587441	Standard
	Sn	118	705.0	11.1	**0.0141**	0.014	98.8	ug/L	497	Standard
	Sb	123	247.5	32.3	**0.0437**	0.016	37.3	ug/L	78	Standard
	Ba	135	3842.5	3.8	**1.6750**	0.079	4.7	ug/L	24	Standard
	Ce	140	6778.2	3.5				ug/L	25	Standard
>	**Tb**	159	1017174.6	0.9				ug/L	939701	Standard
	Ho	165	125.0	20.8				ug/L	10	Standard
	Tl	203	149.7	88.3	**0.0148**	0.015	98.5	ug/L	4	Standard
	Tl	205	133.3	80.1	**0.0172**	0.014	79.0	ug/L	2	Standard
	Pb	206	748.0	16.8	**0.0568**	0.021	37.3	ug/L	240	Standard
	Pb	207	613.7	10.2	**0.0482**	0.012	24.5	ug/L	198	Standard
	Pb	208	2588.8	11.0	**0.0521**	0.014	26.1	ug/L	883	Standard
	U	238	255.3	28.9	**0.0295**	0.010	33.0	ug/L	3	Standard
>	**Bi**	209	573070.1	0.3				ug/L	523984	Standard

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:30:11
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.8927**	30.982	173.2	mg/L	0	Standard
Mg	24	208.3	5.5	**0.1886**	0.017	8.9	mg/L	37	Standard
K	39	35.0	14.3	**-0.0411**	0.011	27.7	mg/L	38	Standard
Ca	43	31.7	18.2	**-0.6344**	0.363	57.2	mg/L	10	Standard
Fe	54	281.0	17.0	**0.0865**	0.040	45.9	mg/L	179	Standard
Fe	57	258.3	4.5	**-0.2461**	0.039	15.8	mg/L	335	Standard
> **Sc-1**	45	64717.9	1.3				mg/L	61029	Standard
Cl	35	29516.1	1.1				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2206.8	1.5				ug/L	1547	Standard
P	31	18893.4	3.4				ug/L	15286	Standard
S	34	2140.2	7.5				ug/L	1572	Standard
Sr	88	85.0	23.5				ug/L	73	Standard
C	12	63.3	39.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	226.9	17.7				mg/L	6	Standard
Ho-1	165	125.0	20.8				mg/L	10	Standard
Er	166	133.3	38.5				mg/L	13	Standard
I	127	4113.9	27.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		128.386	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.643	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:30:11
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	111.061
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	109.368
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606128101
Sample Date/Time: Wednesday, June 29, 2016 17:31:05
Number of Replicates: 3
Autosampler Position: 233
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95248.6	2.8				ug/L	74699	Standard
	Be	9	20.0	25.0	**0.0254**	0.004	15.8	ug/L	8	Standard
	Al	27	13040665.4	1.7	**69.5953**	1.139	1.6	ug/L	1493	Standard
	Sc	45	60164.2	2.0				ug/L	61029	Standard
	Ti	47	192.7	16.8	**0.4120**	0.087	21.1	ug/L	35	Standard
	V	51	8357.2	4.3	**0.9387**	0.048	5.1	ug/L	1435	Standard
	Cr	52	21860.8	2.2	**2.0003**	0.082	4.1	ug/L	7635	Standard
	Cr	53	28106.1	2.4	**30.8448**	0.828	2.7	ug/L	363	Standard
	Mn	55	35106.1	13.3	**5.0193**	0.674	13.4	ug/L	747	Standard
	Co	59	6553.8	6.0	**0.9799**	0.056	5.7	ug/L	234	Standard
	Ni	60	21271.9	1.7	**11.5228**	0.221	1.9	ug/L	215	Standard
	Cu	65	394730.4	1.5	**217.3087**	4.048	1.9	ug/L	104	Standard
	Zn	66	2800.3	3.3	**2.1735**	0.095	4.3	ug/L	363	Standard
>	**Ge**	72	483031.9	0.6				ug/L	495142	Standard
	As	75	2746.8	5.9	**2.5735**	0.152	5.9	ug/L	-50	Standard
	Se	82	892.7	0.7	**8.1998**	0.096	1.2	ug/L	14	Standard
	Se-1	77	2323.2	2.5	**31.9871**	0.998	3.1	ug/L	42	Standard
>	**Ga**	71	476.7	6.4				mg/L	42	Standard
	Rb	85	346107.5	1.7				ug/L	17	Standard
	Y	89	415017.7	1.3				ug/L	417098	Standard
>	**Rh**	103	195.0	26.0				ug/L	13	Standard
	Mo	98	2991.9	3.1	**1.1870**	0.038	3.2	ug/L	11	Standard
	Ag	107	283.3	10.3	**0.0333**	0.004	13.4	ug/L	68	Standard
	Cd	111	47.5	106.4	**0.0162**	0.025	157.7	mg/L	2	Standard
	Cd	114	106.7	47.6	**0.0200**	0.010	52.0	ug/L	12	Standard
>	**In**	115	552878.2	0.7				ug/L	587441	Standard
	Sn	118	3292.0	3.2	**0.5041**	0.021	4.2	ug/L	497	Standard
	Sb	123	2817.9	0.9	**0.6371**	0.010	1.5	ug/L	78	Standard
	Ba	135	115236.2	0.6	**59.8088**	0.708	1.2	ug/L	24	Standard
	Ce	140	265.0	80.7				ug/L	25	Standard
>	**Tb**	159	913633.5	1.5				ug/L	939701	Standard
	Ho	165	16.7	45.8				ug/L	10	Standard
	Tl	203	132.0	58.6	**0.0163**	0.011	65.5	ug/L	4	Standard
	Tl	205	136.7	39.5	**0.0217**	0.009	39.6	ug/L	2	Standard
	Pb	206	460.3	26.2	**0.0258**	0.026	100.0	ug/L	240	Standard
	Pb	207	366.3	38.5	**0.0169**	0.034	199.7	ug/L	198	Standard
	Pb	208	1427.4	18.1	**0.0116**	0.016	136.8	ug/L	883	Standard
	U	238	88.7	86.2	**0.0101**	0.013	124.4	ug/L	3	Standard
>	**Bi**	209	465790.3	0.8				ug/L	523984	Standard

Sample ID: L1606128101
Report Date/Time: Wednesday, June 29, 2016 17:33:22
Page 1

Approved: July 04, 2016

Na	23	8.3	69.3	**97.9562**	69.038	70.5	mg/L	0	Standard
Mg	24	359246.1	3.7	**470.2454**	9.358	2.0	mg/L	37	Standard
K	39	24917.1	2.4	**55.7677**	1.861	3.3	mg/L	38	Standard
Ca	43	1330.1	6.9	**82.7516**	6.716	8.1	mg/L	10	Standard
Fe	54	200.7	20.7	**0.0339**	0.032	94.4	mg/L	179	Standard
Fe	57	1015.0	10.0	**2.4309**	0.415	17.1	mg/L	335	Standard
> **Sc-1**	45	60164.2	2.0				mg/L	61029	Standard
Cl	35	38169.4	0.5				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	223802.3	3.7				ug/L	1547	Standard
P	31	18462.9	1.8				ug/L	15286	Standard
S	34	2156.8	3.7				ug/L	1572	Standard
Sr	88	483.3	16.2				ug/L	73	Standard
C	12	163.3	15.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	24.7				mg/L	3	Standard
Dy	164	32.2	122.0				mg/L	6	Standard
Ho-1	165	16.7	45.8				mg/L	10	Standard
Er	166	23.3	99.0				mg/L	13	Standard
I	127	949274.2	9.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		127.510	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.554	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606128101
Report Date/Time: Wednesday, June 29, 2016 17:33:22
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.116
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		88.894
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Cu 65 Upper, S, EEE	Cu	65	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 17:34:19
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	89879.3	3.0				ug/L	74699	Standard
	Be	9	54419.2	2.1	**51.4171**	2.152	4.2	ug/L	8	Standard
	Al	27	8256861.3	1.5	**46.7135**	1.756	3.8	ug/L	1493	Standard
	Sc	45	65086.2	2.0				ug/L	61029	Standard
	Ti	47	41554.9	1.2	**98.5544**	1.167	1.2	ug/L	35	Standard
	V	51	424002.3	1.3	**50.3253**	0.743	1.5	ug/L	1435	Standard
	Cr	52	412948.6	1.6	**49.1234**	1.130	2.3	ug/L	7635	Standard
	Cr	53	52872.1	1.4	**51.1473**	0.096	0.2	ug/L	363	Standard
	Mn	55	374015.3	1.0	**48.1957**	0.700	1.5	ug/L	747	Standard
	Co	59	377049.8	1.0	**51.3472**	0.802	1.6	ug/L	234	Standard
	Ni	60	101943.5	0.6	**48.8763**	0.675	1.4	ug/L	215	Standard
	Cu	65	101706.9	0.4	**49.0730**	0.490	1.0	ug/L	104	Standard
	Zn	66	60100.6	1.2	**49.0556**	0.165	0.3	ug/L	363	Standard
>	Ge	72	550256.1	1.3				ug/L	495142	Standard
	As	75	60702.3	2.0	**48.8234**	0.606	1.2	ug/L	-50	Standard
	Se	82	6059.9	1.5	**49.5453**	0.269	0.5	ug/L	14	Standard
	Se-1	77	4264.3	0.7	**51.9185**	0.539	1.0	ug/L	42	Standard
>	Ga	71	80.0	16.5				mg/L	42	Standard
	Rb	85	943.4	9.9				ug/L	17	Standard
	Y	89	456697.4	1.9				ug/L	417098	Standard
>	Rh	103	31.7	32.9				ug/L	13	Standard
	Mo	98	286840.3	1.0	**98.7855**	2.014	2.0	ug/L	11	Standard
	Ag	107	350242.5	1.1	**49.3132**	0.224	0.5	ug/L	68	Standard
	Cd	111	110565.3	0.7	**48.5306**	0.244	0.5	mg/L	2	Standard
	Cd	114	270182.5	3.7	**48.3546**	1.830	3.8	ug/L	12	Standard
>	In	115	640022.5	1.2				ug/L	587441	Standard
	Sn	118	309740.9	1.1	**48.5773**	0.779	1.6	ug/L	497	Standard
	Sb	123	247251.7	1.0	**48.6061**	0.525	1.1	ug/L	78	Standard
	Ba	135	103942.0	1.5	**46.5980**	0.695	1.5	ug/L	24	Standard
	Ce	140	81.7	18.7				ug/L	25	Standard
>	Tb	159	1011086.0	0.8				ug/L	939701	Standard
	Ho	165	13.3	94.4				ug/L	10	Standard
	Tl	203	438780.7	1.5	**49.1134**	0.474	1.0	ug/L	4	Standard
	Tl	205	398617.2	2.3	**51.4045**	0.522	1.0	ug/L	2	Standard
	Pb	206	284136.3	1.1	**49.0127**	0.824	1.7	ug/L	240	Standard
	Pb	207	252190.2	0.9	**48.9661**	0.976	2.0	ug/L	198	Standard
	Pb	208	1023481.0	1.1	**50.2926**	0.682	1.4	ug/L	883	Standard
	U	238	380510.7	2.1	**50.9985**	0.305	0.6	ug/L	3	Standard
>	Bi	209	567945.0	2.2				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:36:36
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.3524**	31.779	173.2	mg/L	0	Standard
Mg	24	4450.7	2.4	**5.3249**	0.232	4.3	mg/L	37	Standard
K	39	2040.1	4.0	**4.1161**	0.223	5.4	mg/L	38	Standard
Ca	43	101.7	19.9	**3.5019**	1.219	34.8	mg/L	10	Standard
Fe	54	6522.9	1.9	**5.0481**	0.108	2.1	mg/L	179	Standard
Fe	57	1916.8	5.3	**5.0325**	0.229	4.6	mg/L	335	Standard
Sc-1	45	65086.2	2.0				mg/L	61029	Standard
Cl	35	28870.1	1.5				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	3490.4	22.1				ug/L	1547	Standard
P	31	20260.2	3.0				ug/L	15286	Standard
S	34	2420.2	2.6				ug/L	1572	Standard
Sr	88	90.0	19.2				ug/L	73	Standard
C	12	86.7	6.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	9.5	5.0				mg/L	6	Standard
Ho-1	165	13.3	94.4				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	23970.8	64.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	102.834		
Al	27	93.427		
Sc	45			
Ti	47	98.554		
V	51	100.651		
Cr	52	98.247		
Cr	53			
Mn	55	96.391		
Co	59	102.694		
Ni	60	97.753		
Cu	65	98.146		
Zn	66	98.111		
Ge	72		111.131	
As	75	97.647		
Se	82	99.091		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:36:36
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	98.786		
Ag	107	98.626		
Cd	111	97.061		
Cd	114			
> In	115			108.951
Sn	118	97.155		
Sb	123	97.212		
Ba	135	93.196		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	98.227		
Tl	205			
Pb	206			
Pb	207			
Pb	208	100.585		
U	238	101.997		
> Bi	209			108.390
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:36:36
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 17:37:30
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90192.8	3.2				ug/L	74699	Standard
	Be	9	55.0	102.4	**0.0584**	0.051	86.9	ug/L	8	Standard
	Al	27	41697.2	123.7	**0.2160**	0.280	129.5	ug/L	1493	Standard
	Sc	45	66865.6	0.3				ug/L	61029	Standard
	Ti	47	70.3	63.0	**0.0618**	0.107	173.0	ug/L	35	Standard
	V	51	1545.5	1.7	**-0.0068**	0.004	62.6	ug/L	1435	Standard
	Cr	52	7551.9	5.0	**-0.0837**	0.044	52.6	ug/L	7635	Standard
	Cr	53	1056.7	15.9	**0.7243**	0.162	22.4	ug/L	363	Standard
	Mn	55	2100.5	33.7	**0.1299**	0.092	71.2	ug/L	747	Standard
	Co	59	307.3	14.8	**0.0056**	0.006	108.6	ug/L	234	Standard
	Ni	60	288.7	41.6	**0.0182**	0.059	321.7	ug/L	215	Standard
	Cu	65	326.7	64.3	**0.0580**	0.103	178.3	ug/L	104	Standard
	Zn	66	465.0	14.2	**-0.0620**	0.054	86.6	ug/L	363	Standard
>	**Ge**	72	536712.4	0.6				ug/L	495142	Standard
	As	75	-14.2	293.5	**0.0478**	0.034	71.9	ug/L	-50	Standard
	Se	82	26.6	3.7	**0.0859**	0.008	9.9	ug/L	14	Standard
	Se-1	77	71.3	25.4	**0.2781**	0.225	81.0	ug/L	42	Standard
>	**Ga**	71	51.7	22.3				mg/L	42	Standard
	Rb	85	111.7	73.6				ug/L	17	Standard
	Y	89	476031.3	1.1				ug/L	417098	Standard
>	**Rh**	103	26.7	47.2				ug/L	13	Standard
	Mo	98	310.3	31.4	**0.0963**	0.033	33.9	ug/L	11	Standard
	Ag	107	196.3	41.5	**0.0135**	0.011	82.4	ug/L	68	Standard
	Cd	111	44.4	76.9	**0.0108**	0.014	134.5	mg/L	2	Standard
	Cd	114	106.2	73.2	**0.0161**	0.013	83.2	ug/L	12	Standard
>	**In**	115	671001.8	0.7				ug/L	587441	Standard
	Sn	118	950.0	11.1	**0.0476**	0.015	31.2	ug/L	497	Standard
	Sb	123	463.6	17.0	**0.0827**	0.014	17.2	ug/L	78	Standard
	Ba	135	113.3	53.3	**0.0326**	0.026	80.4	ug/L	24	Standard
	Ce	140	100.0	52.9				ug/L	25	Standard
>	**Tb**	159	1042844.6	0.4				ug/L	939701	Standard
	Ho	165	21.7	53.3				ug/L	10	Standard
	Tl	203	169.7	59.1	**0.0166**	0.011	64.7	ug/L	4	Standard
	Tl	205	136.7	25.7	**0.0173**	0.004	24.8	ug/L	2	Standard
	Pb	206	350.0	21.1	**-0.0125**	0.012	95.5	ug/L	240	Standard
	Pb	207	292.7	21.7	**-0.0148**	0.012	78.6	ug/L	198	Standard
	Pb	208	1272.0	10.4	**-0.0133**	0.006	44.8	ug/L	883	Standard
	U	238	124.3	70.3	**0.0117**	0.011	96.0	ug/L	3	Standard
>	**Bi**	209	584508.5	0.7				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:39:47
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	165.0	89.4	**0.1290**	0.173	134.2	mg/L	37	Standard
K	39	35.0	14.3	**-0.0435**	0.010	22.9	mg/L	38	Standard
Ca	43	33.3	22.9	**-0.6014**	0.442	73.5	mg/L	10	Standard
Fe	54	222.3	16.4	**0.0337**	0.028	82.1	mg/L	179	Standard
Fe	57	293.3	1.0	**-0.1642**	0.011	6.9	mg/L	335	Standard
> **Sc-1**	45	66865.6	0.3				mg/L	61029	Standard
Cl	35	28993.0	1.1				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	2536.9	15.9				ug/L	1547	Standard
P	31	21593.7	3.8				ug/L	15286	Standard
S	34	2441.9	4.5				ug/L	1572	Standard
Sr	88	71.7	39.7				ug/L	73	Standard
C	12	43.3	53.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	2.2	286.4				mg/L	6	Standard
Ho-1	165	21.7	53.3				mg/L	10	Standard
Er	166	23.3	65.5				mg/L	13	Standard
I	127	6326.4	22.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.396	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:39:47
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.225
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.551
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:39:47
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129403
Sample Date/Time: Wednesday, June 29, 2016 17:40:43
Number of Replicates: 3
Autosampler Position: 234
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	91138.4	2.1				ug/L	74699	Standard
	Be	9	18.3	31.5	**0.0246**	0.005	20.1	ug/L	8	Standard
	Al	27	19782.9	2.5	**0.0970**	0.003	3.0	ug/L	1493	Standard
	Sc	45	69032.0	3.0				ug/L	61029	Standard
	Ti	47	54.0	3.2	**0.0194**	0.002	12.6	ug/L	35	Standard
	V	51	1966.3	6.5	**0.0390**	0.011	28.1	ug/L	1435	Standard
	Cr	52	10805.9	1.1	**0.2915**	0.043	14.7	ug/L	7635	Standard
	Cr	53	2143.5	1.8	**1.7623**	0.077	4.4	ug/L	363	Standard
	Mn	55	3844.5	2.2	**0.3500**	0.018	5.0	ug/L	747	Standard
	Co	59	327.3	8.7	**0.0074**	0.003	41.0	ug/L	234	Standard
	Ni	60	356.0	5.4	**0.0475**	0.008	16.5	ug/L	215	Standard
	Cu	65	696.0	2.6	**0.2337**	0.014	6.0	ug/L	104	Standard
	Zn	66	2205.8	3.0	**1.3661**	0.024	1.8	ug/L	363	Standard
>	**Ge**	72	549152.6	2.2				ug/L	495142	Standard
	As	75	-29.6	63.4	**0.0356**	0.015	42.9	ug/L	-50	Standard
	Se	82	23.9	20.2	**0.0586**	0.040	68.9	ug/L	14	Standard
	Se-1	77	94.7	7.0	**0.5478**	0.105	19.2	ug/L	42	Standard
>	**Ga**	71	51.7	14.8				mg/L	42	Standard
	Rb	85	138.3	25.6				ug/L	17	Standard
	Y	89	465979.4	3.6				ug/L	417098	Standard
>	**Rh**	103	13.3	21.7				ug/L	13	Standard
	Mo	98	5139.1	1.4	**1.7220**	0.056	3.3	ug/L	11	Standard
	Ag	107	110.0	6.9	**0.0022**	0.001	48.4	ug/L	68	Standard
	Cd	111	15.0	26.6	**-0.0015**	0.002	117.9	mg/L	2	Standard
	Cd	114	55.2	20.3	**0.0075**	0.002	25.2	ug/L	12	Standard
>	**In**	115	655841.8	1.8				ug/L	587441	Standard
	Sn	118	1048.4	10.6	**0.0659**	0.015	22.6	ug/L	497	Standard
	Sb	123	219.8	14.7	**0.0379**	0.005	14.5	ug/L	78	Standard
	Ba	135	307.7	5.8	**0.1187**	0.008	7.0	ug/L	24	Standard
	Ce	140	120.0	27.3				ug/L	25	Standard
>	**Tb**	159	1025647.2	2.3				ug/L	939701	Standard
	Ho	165	10.0	50.0				ug/L	10	Standard
	Tl	203	102.7	26.3	**0.0095**	0.003	29.3	ug/L	4	Standard
	Tl	205	93.3	17.2	**0.0120**	0.002	17.6	ug/L	2	Standard
	Pb	206	348.0	5.5	**-0.0123**	0.003	25.6	ug/L	240	Standard
	Pb	207	291.0	7.0	**-0.0146**	0.003	21.1	ug/L	198	Standard
	Pb	208	1220.0	1.5	**-0.0152**	0.002	10.4	ug/L	883	Standard
	U	238	43.7	3.5	**0.0013**	0.000	20.2	ug/L	3	Standard
>	**Bi**	209	579311.4	1.7				ug/L	523984	Standard

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:43:00
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	70.0	14.3	**0.0151**	0.014	91.8	mg/L	37	Standard
K	39	41.7	36.7	**-0.0323**	0.032	98.5	mg/L	38	Standard
Ca	43	23.3	24.7	**-1.2220**	0.307	25.1	mg/L	10	Standard
Fe	54	180.7	18.1	**-0.0033**	0.021	642.1	mg/L	179	Standard
Fe	57	228.3	18.9	**-0.3899**	0.112	28.8	mg/L	335	Standard
Sc-1	45	69032.0	3.0				mg/L	61029	Standard
Cl	35	29393.1	0.1				ug/L	20195	Standard
Kr	83	3.7	41.7				ug/L	4	Standard
Br	81	2280.2	16.1				ug/L	1547	Standard
P	31	20889.4	5.8				ug/L	15286	Standard
S	34	2440.2	2.2				ug/L	1572	Standard
Sr	88	78.3	9.8				ug/L	73	Standard
C	12	90.0	19.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	5.2	126.6				mg/L	6	Standard
Ho-1	165	10.0	50.0				mg/L	10	Standard
Er	166	30.0	66.7				mg/L	13	Standard
I	127	4447.3	11.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		122.008	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		110.908	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:43:00
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.644
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.559
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129403
Sample Date/Time: Wednesday, June 29, 2016 17:43:55
Number of Replicates: 3
Autosampler Position: 235
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89753.4	1.3				ug/L	74699	Standard
	Be	9	13.3	43.3	**0.0203**	0.006	27.7	ug/L	8	Standard
	Al	27	7761551.3	2.6	**43.9444**	1.176	2.7	ug/L	1493	Standard
	Sc	45	64855.2	1.4				ug/L	61029	Standard
	Ti	47	206.0	13.1	**0.3915**	0.063	16.1	ug/L	35	Standard
	V	51	6651.8	1.7	**0.6159**	0.016	2.5	ug/L	1435	Standard
	Cr	52	10534.3	3.3	**0.2860**	0.045	15.8	ug/L	7635	Standard
	Cr	53	2358.5	8.3	**2.0219**	0.206	10.2	ug/L	363	Standard
	Mn	55	48108.9	1.2	**6.2182**	0.039	0.6	ug/L	747	Standard
	Co	59	779.0	1.7	**0.0714**	0.002	2.2	ug/L	234	Standard
	Ni	60	2066.1	2.9	**0.8928**	0.027	3.0	ug/L	215	Standard
	Cu	65	1897.5	0.8	**0.8356**	0.011	1.4	ug/L	104	Standard
	Zn	66	4161.9	4.0	**3.0555**	0.122	4.0	ug/L	363	Standard
>	**Ge**	72	537411.3	0.6				ug/L	495142	Standard
	As	75	1930.1	2.9	**1.6472**	0.048	2.9	ug/L	-50	Standard
	Se	82	65.9	12.2	**0.4151**	0.071	17.1	ug/L	14	Standard
	Se-1	77	143.0	7.4	**1.1820**	0.141	12.0	ug/L	42	Standard
>	**Ga**	71	281.7	11.8				mg/L	42	Standard
	Rb	85	9322.9	0.6				ug/L	17	Standard
	Y	89	483828.6	0.3				ug/L	417098	Standard
>	**Rh**	103	65.0	35.3				ug/L	13	Standard
	Mo	98	2420.2	2.2	**0.8025**	0.008	1.0	ug/L	11	Standard
	Ag	107	86.7	9.0	**-0.0011**	0.001	98.3	ug/L	68	Standard
	Cd	111	21.1	19.4	**0.0011**	0.002	172.0	mg/L	2	Standard
	Cd	114	57.2	8.0	**0.0078**	0.001	8.8	ug/L	12	Standard
>	**In**	115	659947.0	1.4				ug/L	587441	Standard
	Sn	118	1036.7	2.5	**0.0632**	0.002	2.8	ug/L	497	Standard
	Sb	123	514.2	4.3	**0.0939**	0.005	5.0	ug/L	78	Standard
	Ba	135	42292.6	1.8	**18.3786**	0.354	1.9	ug/L	24	Standard
	Ce	140	906.7	5.0				ug/L	25	Standard
>	**Tb**	159	1004743.5	1.0				ug/L	939701	Standard
	Ho	165	30.0	33.3				ug/L	10	Standard
	Tl	203	321.7	3.2	**0.0348**	0.001	3.5	ug/L	4	Standard
	Tl	205	263.3	19.0	**0.0346**	0.006	17.3	ug/L	2	Standard
	Pb	206	528.0	6.8	**0.0212**	0.005	24.5	ug/L	240	Standard
	Pb	207	478.3	8.2	**0.0244**	0.009	35.8	ug/L	198	Standard
	Pb	208	1973.7	3.1	**0.0244**	0.005	18.6	ug/L	883	Standard
	U	238	903.7	2.1	**0.1184**	0.002	1.8	ug/L	3	Standard
>	**Bi**	209	560080.8	1.7				ug/L	523984	Standard

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:46:12
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	9985.0	3.9	**12.0618**	0.357	3.0	mg/L	37	Standard
K	39	420.0	7.4	**0.7593**	0.061	8.1	mg/L	38	Standard
Ca	43	263.3	2.9	**13.1350**	0.654	5.0	mg/L	10	Standard
Fe	54	210.7	11.6	**0.0297**	0.017	57.3	mg/L	179	Standard
Fe	57	393.3	9.5	**0.1838**	0.119	64.8	mg/L	335	Standard
> **Sc-1**	45	64855.2	1.4				mg/L	61029	Standard
Cl	35	32877.0	2.8				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	8902.6	4.1				ug/L	1547	Standard
P	31	19829.6	4.1				ug/L	15286	Standard
S	34	2240.2	4.2				ug/L	1572	Standard
Sr	88	221.7	14.7				ug/L	73	Standard
C	12	126.7	4.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	64.1	36.5				mg/L	6	Standard
Ho-1	165	30.0	33.3				mg/L	10	Standard
Er	166	53.3	21.7				mg/L	13	Standard
I	127	24610.0	6.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		120.154	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.537	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:46:12
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.343
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.889
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129601
Sample Date/Time: Wednesday, June 29, 2016 17:47:06
Number of Replicates: 3
Autosampler Position: 236
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89414.7	0.8				ug/L	74699	Standard
	Be	9	18.3	41.7	**0.0250**	0.007	28.7	ug/L	8	Standard
	Al	27	159581.9	4.0	**0.8936**	0.029	3.3	ug/L	1493	Standard
	Sc	45	67088.2	1.1				ug/L	61029	Standard
	Ti	47	35.7	1.6	**-0.0268**	0.002	7.8	ug/L	35	Standard
	V	51	2172.9	6.2	**0.0559**	0.017	29.8	ug/L	1435	Standard
	Cr	52	9358.6	1.2	**0.0800**	0.026	32.1	ug/L	7635	Standard
	Cr	53	761.7	7.6	**0.3908**	0.063	16.2	ug/L	363	Standard
	Mn	55	2132.8	0.4	**0.1197**	0.005	4.4	ug/L	747	Standard
	Co	59	232.7	3.6	**-0.0065**	0.001	22.5	ug/L	234	Standard
	Ni	60	263.3	7.3	**-0.0009**	0.011	1213.2	ug/L	215	Standard
	Cu	65	215.7	8.8	**-0.0024**	0.009	395.1	ug/L	104	Standard
	Zn	66	2535.5	4.0	**1.5735**	0.060	3.8	ug/L	363	Standard
>	**Ge**	72	566628.0	1.5				ug/L	495142	Standard
	As	75	0.7	2838.5	**0.0602**	0.015	24.5	ug/L	-50	Standard
	Se	82	16.5	13.6	**-0.0062**	0.020	325.1	ug/L	14	Standard
	Se-1	77	57.7	30.4	**0.0671**	0.205	305.9	ug/L	42	Standard
>	**Ga**	71	65.0	35.3				mg/L	42	Standard
	Rb	85	218.3	8.7				ug/L	17	Standard
	Y	89	474347.0	1.6				ug/L	417098	Standard
>	**Rh**	103	16.7	45.8				ug/L	13	Standard
	Mo	98	108.8	9.6	**0.0309**	0.003	10.7	ug/L	11	Standard
	Ag	107	170.0	11.0	**0.0105**	0.003	26.1	ug/L	68	Standard
	Cd	111	6.5	32.0	**-0.0051**	0.001	17.0	mg/L	2	Standard
	Cd	114	10.6	76.1	**-0.0002**	0.001	576.3	ug/L	12	Standard
>	**In**	115	655393.1	1.0				ug/L	587441	Standard
	Sn	118	536.7	6.3	**-0.0123**	0.006	49.5	ug/L	497	Standard
	Sb	123	179.2	10.0	**0.0302**	0.004	12.2	ug/L	78	Standard
	Ba	135	901.0	1.9	**0.3786**	0.012	3.1	ug/L	24	Standard
	Ce	140	38.3	52.7				ug/L	25	Standard
>	**Tb**	159	1019773.4	1.2				ug/L	939701	Standard
	Ho	165	10.0	50.0				ug/L	10	Standard
	Tl	203	34.7	13.6	**0.0021**	0.001	27.9	ug/L	4	Standard
	Tl	205	33.3	52.7	**0.0044**	0.002	51.3	ug/L	2	Standard
	Pb	206	333.0	5.0	**-0.0144**	0.003	23.5	ug/L	240	Standard
	Pb	207	282.3	6.2	**-0.0157**	0.004	25.6	ug/L	198	Standard
	Pb	208	1187.0	2.0	**-0.0164**	0.002	10.6	ug/L	883	Standard
	U	238	27.0	25.9	**-0.0009**	0.001	108.4	ug/L	3	Standard
>	**Bi**	209	574894.3	1.4				ug/L	523984	Standard

Sample ID: L1606129601
Report Date/Time: Wednesday, June 29, 2016 17:49:23
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.3542**	30.050	173.2	mg/L	0	Standard
Mg	24	255.0	18.0	**0.2345**	0.054	23.2	mg/L	37	Standard
K	39	41.7	36.7	**-0.0306**	0.030	98.0	mg/L	38	Standard
Ca	43	31.7	18.2	**-0.7022**	0.343	48.8	mg/L	10	Standard
Fe	54	182.6	4.3	**0.0025**	0.004	177.8	mg/L	179	Standard
Fe	57	238.3	19.9	**-0.3363**	0.155	46.0	mg/L	335	Standard
> **Sc-1**	45	67088.2	1.1				mg/L	61029	Standard
Cl	35	32679.2	0.8				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	2083.5	4.4				ug/L	1547	Standard
P	31	19699.5	2.2				ug/L	15286	Standard
S	34	2248.5	3.2				ug/L	1572	Standard
Sr	88	70.0	56.7				ug/L	73	Standard
C	12	60.0	76.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	42.4	37.5				mg/L	6	Standard
Ho-1	165	10.0	50.0				mg/L	10	Standard
Er	166	20.0	86.6				mg/L	13	Standard
I	127	3452.1	27.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		119.700	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.437	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129601
Report Date/Time: Wednesday, June 29, 2016 17:49:23
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.567
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.716
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Wednesday, June 29, 2016 17:50:19
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88962.1	0.9				ug/L	74699	Standard
	Be	9	13.3	43.3	**0.0204**	0.006	27.5	ug/L	8	Standard
	Al	27	8687475.6	2.0	**49.6312**	1.401	2.8	ug/L	1493	Standard
	Sc	45	67439.8	1.2				ug/L	61029	Standard
	Ti	47	49495.1	1.8	**118.6714**	2.776	2.3	ug/L	35	Standard
	V	51	2497.1	11.5	**0.1050**	0.033	31.3	ug/L	1435	Standard
	Cr	52	8893.6	0.7	**0.0677**	0.012	17.4	ug/L	7635	Standard
	Cr	53	8300.6	3.8	**7.8359**	0.175	2.2	ug/L	363	Standard
	Mn	55	1506.4	1.9	**0.0487**	0.003	6.1	ug/L	747	Standard
	Co	59	411.3	2.2	**0.0194**	0.002	11.4	ug/L	234	Standard
	Ni	60	986.0	7.4	**0.3548**	0.029	8.2	ug/L	215	Standard
	Cu	65	754.7	3.5	**0.2650**	0.009	3.4	ug/L	104	Standard
	Zn	66	1997.5	3.6	**1.2085**	0.053	4.4	ug/L	363	Standard
>	**Ge**	72	544457.7	1.7				ug/L	495142	Standard
	As	75	-43.8	164.6	**0.0238**	0.058	245.3	ug/L	-50	Standard
	Se	82	18.8	19.8	**0.0181**	0.034	186.0	ug/L	14	Standard
	Se-1	77	455.0	4.4	**5.0437**	0.245	4.9	ug/L	42	Standard
>	**Ga**	71	186.7	17.4				mg/L	42	Standard
	Rb	85	1130.0	9.7				ug/L	17	Standard
	Y	89	464215.3	1.7				ug/L	417098	Standard
>	**Rh**	103	10.0	100.0				ug/L	13	Standard
	Mo	98	300673.2	0.9	**102.5347**	1.457	1.4	ug/L	11	Standard
	Ag	107	252.3	6.3	**0.0223**	0.003	12.6	ug/L	68	Standard
	Cd	111	-194.8	14.0	**-0.0926**	0.012	12.9	mg/L	2	Standard
	Cd	114	577.1	3.2	**0.1002**	0.004	4.2	ug/L	12	Standard
>	**In**	115	646355.5	1.8				ug/L	587441	Standard
	Sn	118	791.7	5.4	**0.0285**	0.006	21.5	ug/L	497	Standard
	Sb	123	195.6	12.5	**0.0338**	0.004	11.9	ug/L	78	Standard
	Ba	135	712.0	4.0	**0.3001**	0.008	2.7	ug/L	24	Standard
	Ce	140	246.7	7.1				ug/L	25	Standard
>	**Tb**	159	1024202.6	1.1				ug/L	939701	Standard
	Ho	165	23.3	61.9				ug/L	10	Standard
	Tl	203	122.3	9.8	**0.0119**	0.001	11.8	ug/L	4	Standard
	Tl	205	90.0	11.1	**0.0118**	0.001	10.5	ug/L	2	Standard
	Pb	206	533.7	2.0	**0.0208**	0.002	10.5	ug/L	240	Standard
	Pb	207	414.7	3.2	**0.0105**	0.002	21.4	ug/L	198	Standard
	Pb	208	1800.4	3.2	**0.0144**	0.003	21.4	ug/L	883	Standard
	U	238	18.7	20.3	**-0.0019**	0.001	26.8	ug/L	3	Standard
>	**Bi**	209	568876.5	0.5				ug/L	523984	Standard

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 17:52:36
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	11507.7	5.3	**13.3841**	0.868	6.5	mg/L	37	Standard
K	39	2163.5	5.0	**4.2129**	0.203	4.8	mg/L	38	Standard
Ca	43	273.3	20.1	**13.1286**	3.349	25.5	mg/L	10	Standard
Fe	54	17120.6	2.5	**12.9991**	0.385	3.0	mg/L	179	Standard
Fe	57	4470.7	3.8	**12.6745**	0.387	3.1	mg/L	335	Standard
> **Sc-1**	45	67439.8	1.2				mg/L	61029	Standard
Cl	35	28777.3	2.0				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	2086.8	10.9				ug/L	1547	Standard
P	31	20233.5	1.2				ug/L	15286	Standard
S	34	2398.5	7.6				ug/L	1572	Standard
Sr	88	63.3	35.6				ug/L	73	Standard
C	12	173.3	8.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	80.0	69.6				mg/L	3	Standard
Dy	164	11.6	52.9				mg/L	6	Standard
Ho-1	165	23.3	61.9				mg/L	10	Standard
Er	166	36.7	56.8				mg/L	13	Standard
I	127	2830.3	21.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27	0.993		
Sc	45			
Ti	47	118.671		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.960	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 17:52:36
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.535		
Ag	107			
Cd	111			
Cd	114			
In	115			110.029
Sn	118			
Sb	123			
Ba	135			
Ce	140			
Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
Bi	209			108.567
Na	23	0.040		
Mg	24	267.683		
K	39	84.259		
Ca	43	87.524		
Fe	54	103.992		
Fe	57	101.396		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Na	23	
QC Std 4	Mg	24	

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 17:52:36
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Wednesday, June 29, 2016 17:53:31
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	88668.7	1.6				ug/L	74699	Standard
	Be	9	115505.6	3.3	**110.5278**	2.817	2.5	ug/L	8	Standard
	Al	27	8938177.8	2.5	**51.2202**	0.471	0.9	ug/L	1493	Standard
	Sc	45	65087.9	3.1				ug/L	61029	Standard
	Ti	47	50197.4	1.2	**122.5394**	1.491	1.2	ug/L	35	Standard
	V	51	859522.7	1.3	**105.1949**	1.380	1.3	ug/L	1435	Standard
	Cr	52	831098.2	2.0	**102.8267**	1.943	1.9	ug/L	7635	Standard
	Cr	53	106911.0	2.3	**106.7884**	2.096	2.0	ug/L	363	Standard
	Mn	55	780341.4	1.6	**103.6433**	1.453	1.4	ug/L	747	Standard
	Co	59	785536.0	0.5	**110.1234**	0.547	0.5	ug/L	234	Standard
	Ni	60	203753.2	2.0	**100.6522**	1.808	1.8	ug/L	215	Standard
	Cu	65	201732.1	0.8	**100.2703**	0.833	0.8	ug/L	104	Standard
	Zn	66	122540.9	0.7	**103.4323**	0.376	0.4	ug/L	363	Standard
>	Ge	72	534677.6	0.4				ug/L	495142	Standard
	As	75	124456.8	0.2	**102.9556**	0.389	0.4	ug/L	-50	Standard
	Se	82	12596.7	0.4	**106.1488**	0.433	0.4	ug/L	14	Standard
	Se-1	77	8794.6	0.8	**110.8893**	1.382	1.2	ug/L	42	Standard
>	Ga	71	163.3	10.8				mg/L	42	Standard
	Rb	85	966.7	4.8				ug/L	17	Standard
	Y	89	476525.6	1.7				ug/L	417098	Standard
>	Rh	103	56.7	53.9				ug/L	13	Standard
	Mo	98	309227.9	0.6	**104.8818**	1.132	1.1	ug/L	11	Standard
	Ag	107	592523.9	2.1	**82.1821**	1.932	2.4	ug/L	68	Standard
	Cd	111	228625.7	1.0	**98.8467**	0.904	0.9	mg/L	2	Standard
	Cd	114	551600.1	1.5	**97.2335**	1.523	1.6	ug/L	12	Standard
>	In	115	649797.8	0.6				ug/L	587441	Standard
	Sn	118	1576.7	4.2	**0.1493**	0.011	7.5	ug/L	497	Standard
	Sb	123	495027.3	1.9	**95.8585**	2.358	2.5	ug/L	78	Standard
	Ba	135	210763.3	1.3	**93.0830**	1.687	1.8	ug/L	24	Standard
	Ce	140	70.0	24.7				ug/L	25	Standard
>	Tb	159	1005734.4	0.5				ug/L	939701	Standard
	Ho	165	11.7	49.5				ug/L	10	Standard
	Tl	203	893894.1	1.2	**101.2389**	0.317	0.3	ug/L	4	Standard
	Tl	205	815034.2	0.6	**106.3608**	0.654	0.6	ug/L	2	Standard
	Pb	206	583528.7	0.9	**101.9143**	0.243	0.2	ug/L	240	Standard
	Pb	207	499626.1	1.2	**98.2110**	0.372	0.4	ug/L	198	Standard
	Pb	208	2105185.2	0.8	**104.7451**	0.431	0.4	ug/L	883	Standard
	U	238	779429.7	1.7	**105.7078**	0.634	0.6	ug/L	3	Standard
>	Bi	209	561247.5	1.2				ug/L	523984	Standard

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.1880**	31.494	173.2	mg/L	0	Standard
Mg	24	12009.8	3.1	**14.4710**	0.020	0.1	mg/L	37	Standard
K	39	2150.2	4.2	**4.3456**	0.256	5.9	mg/L	38	Standard
Ca	43	251.7	12.1	**12.4083**	2.087	16.8	mg/L	10	Standard
Fe	54	17836.2	2.6	**14.0504**	0.592	4.2	mg/L	179	Standard
Fe	57	4537.3	4.4	**13.3981**	0.845	6.3	mg/L	335	Standard
> **Sc-1**	45	65087.9	3.1				mg/L	61029	Standard
Cl	35	27559.7	2.9				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	1986.8	4.8				ug/L	1547	Standard
P	31	19052.0	2.9				ug/L	15286	Standard
S	34	2286.8	6.3				ug/L	1572	Standard
Sr	88	50.0	10.0				ug/L	73	Standard
C	12	120.0	54.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	6.0	200.7				mg/L	6	Standard
Ho-1	165	11.7	49.5				mg/L	10	Standard
Er	166	13.3	86.6				mg/L	13	Standard
I	127	2065.1	25.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	110.528		
Al	27	1.024		
Sc	45			
Ti	47	122.539		
V	51	105.195		
Cr	52	102.827		
Cr	53			
Mn	55	103.643		
Co	59	110.123		
Ni	60	100.652		
Cu	65	100.270		
Zn	66	103.432		
> Ge	72		107.985	
As	75	102.956		
Se	82	106.149		
Se-1	77			
> Ga	71			

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	104.882		
Ag	107	82.182		
Cd	111	98.847		
Cd	114			
> In	115		110.615	
Sn	118			
Sb	123	95.858		
Ba	135	93.083		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.239		
Tl	205			
Pb	206			
Pb	207			
Pb	208	104.745		
U	238	105.708		
> Bi	209		107.111	
Na	23	145.504		
Mg	24	289.419		
K	39	86.911		
Ca	43	82.722		
Fe	54	112.403		
Fe	57	107.185		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Ti	47	
QC Std 5	Na	23	

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 3

Approved: July 04, 2016

QC Std 5	Mg	24

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 17:56:44
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88001.7	3.8				ug/L	74699	Standard
	Be	9	58491.1	3.4	**56.4121**	0.807	1.4	ug/L	8	Standard
	Al	27	8741223.1	1.9	**50.5126**	1.685	3.3	ug/L	1493	Standard
	Sc	45	66566.0	2.9				ug/L	61029	Standard
	Ti	47	44065.3	0.6	**105.8464**	2.393	2.3	ug/L	35	Standard
	V	51	439728.9	0.9	**52.8674**	1.388	2.6	ug/L	1435	Standard
	Cr	52	434693.1	0.1	**52.4281**	0.919	1.8	ug/L	7635	Standard
	Cr	53	53735.1	1.4	**52.6541**	1.206	2.3	ug/L	363	Standard
	Mn	55	394255.2	1.0	**51.4522**	0.719	1.4	ug/L	747	Standard
	Co	59	403494.7	0.9	**55.6456**	1.136	2.0	ug/L	234	Standard
	Ni	60	106493.3	0.2	**51.7076**	0.770	1.5	ug/L	215	Standard
	Cu	65	105229.8	0.3	**51.4208**	1.011	2.0	ug/L	104	Standard
	Zn	66	61955.4	1.5	**51.2322**	1.156	2.3	ug/L	363	Standard
>	**Ge**	72	543443.2	1.7				ug/L	495142	Standard
	As	75	62757.8	1.3	**51.1208**	1.290	2.5	ug/L	-50	Standard
	Se	82	6411.0	0.6	**53.0945**	1.065	2.0	ug/L	14	Standard
	Se-1	77	4401.6	1.7	**54.2995**	1.443	2.7	ug/L	42	Standard
>	**Ga**	71	61.7	12.4				mg/L	42	Standard
	Rb	85	943.4	4.4				ug/L	17	Standard
	Y	89	472814.1	2.0				ug/L	417098	Standard
>	**Rh**	103	31.7	36.5				ug/L	13	Standard
	Mo	98	301859.8	0.5	**101.5547**	1.508	1.5	ug/L	11	Standard
	Ag	107	366041.5	1.2	**50.3642**	1.561	3.1	ug/L	68	Standard
	Cd	111	117544.4	1.3	**50.4048**	0.748	1.5	mg/L	2	Standard
	Cd	114	286950.2	2.0	**50.1688**	0.955	1.9	ug/L	12	Standard
>	**In**	115	655193.6	1.9				ug/L	587441	Standard
	Sn	118	333959.0	1.2	**51.1818**	1.510	3.0	ug/L	497	Standard
	Sb	123	261510.0	1.1	**50.2357**	1.515	3.0	ug/L	78	Standard
	Ba	135	110572.3	0.3	**48.4318**	0.852	1.8	ug/L	24	Standard
	Ce	140	81.7	46.0				ug/L	25	Standard
>	**Tb**	159	1009652.9	1.9				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	466012.5	0.6	**52.5145**	1.216	2.3	ug/L	4	Standard
	Tl	205	428260.1	0.9	**55.6109**	1.642	3.0	ug/L	2	Standard
	Pb	206	294247.5	0.8	**51.0977**	1.194	2.3	ug/L	240	Standard
	Pb	207	260542.8	0.7	**50.9223**	1.007	2.0	ug/L	198	Standard
	Pb	208	1081629.9	0.8	**53.5152**	1.532	2.9	ug/L	883	Standard
	U	238	403586.7	0.9	**54.4635**	1.402	2.6	ug/L	3	Standard
>	**Bi**	209	564239.9	2.1				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:59:01
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4549.0	2.0	**5.3196**	0.066	1.2	mg/L	37	Standard
K	39	2091.8	2.6	**4.1250**	0.021	0.5	mg/L	38	Standard
Ca	43	106.7	9.8	**3.6589**	0.645	17.6	mg/L	10	Standard
Fe	54	6853.2	4.1	**5.1952**	0.357	6.9	mg/L	179	Standard
Fe	57	1901.8	4.2	**4.8507**	0.081	1.7	mg/L	335	Standard
> **Sc-1**	45	66566.0	2.9				mg/L	61029	Standard
Cl	35	28481.4	2.6				ug/L	20195	Standard
Kr	83	3.3	75.5				ug/L	4	Standard
Br	81	1840.1	18.3				ug/L	1547	Standard
P	31	20427.1	5.4				ug/L	15286	Standard
S	34	2438.5	6.3				ug/L	1572	Standard
Sr	88	80.0	32.5				ug/L	73	Standard
C	12	76.7	7.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	6.5	86.7				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	2150.2	26.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	112.824		
Al	27	101.025		
Sc	45			
Ti	47	105.846		
V	51	105.735		
Cr	52	104.856		
Cr	53			
Mn	55	102.904		
Co	59	111.291		
Ni	60	103.415		
Cu	65	102.842		
Zn	66	102.464		
> Ge	72		109.755	
As	75	102.242		
Se	82	106.189		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:59:01
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.555		
Ag	107	100.728		
Cd	111	100.810		
Cd	114			
In	115		111.533	
Sn	118	102.364		
Sb	123	100.471		
Ba	135	96.864		
Ce	140			
Tb	159			
Ho	165			
Tl	203	105.029		
Tl	205			
Pb	206			
Pb	207			
Pb	208	107.030		
U	238	108.927		
Bi	209		107.683	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:59:01
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 17:59:55
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86353.7	1.4				ug/L	74699	Standard
	Be	9	83.3	67.4	**0.0891**	0.054	60.5	ug/L	8	Standard
	Al	27	42873.5	92.7	**0.2376**	0.231	97.1	ug/L	1493	Standard
	Sc	45	65305.4	1.2				ug/L	61029	Standard
	Ti	47	79.7	51.6	**0.0788**	0.097	123.4	ug/L	35	Standard
	V	51	1461.2	4.4	**-0.0225**	0.008	34.9	ug/L	1435	Standard
	Cr	52	7626.9	6.9	**-0.1029**	0.066	64.1	ug/L	7635	Standard
	Cr	53	760.0	14.8	**0.4057**	0.109	26.8	ug/L	363	Standard
	Mn	55	4546.9	137.2	**0.4388**	0.807	183.9	ug/L	747	Standard
	Co	59	626.0	93.3	**0.0478**	0.080	167.0	ug/L	234	Standard
	Ni	60	409.0	51.9	**0.0715**	0.102	142.7	ug/L	215	Standard
	Cu	65	272.7	44.7	**0.0274**	0.059	215.3	ug/L	104	Standard
	Zn	66	514.0	23.7	**-0.0335**	0.100	300.2	ug/L	363	Standard
>	**Ge**	72	553485.6	1.3				ug/L	495142	Standard
	As	75	-6.7	474.8	**0.0545**	0.025	45.9	ug/L	-50	Standard
	Se	82	20.4	10.3	**0.0278**	0.015	55.0	ug/L	14	Standard
	Se-1	77	66.3	4.4	**0.1906**	0.046	24.0	ug/L	42	Standard
>	**Ga**	71	50.0	10.0				mg/L	42	Standard
	Rb	85	45.0	48.4				ug/L	17	Standard
	Y	89	474090.9	1.7				ug/L	417098	Standard
>	**Rh**	103	18.3	41.7				ug/L	13	Standard
	Mo	98	533.3	54.8	**0.1739**	0.097	56.0	ug/L	11	Standard
	Ag	107	205.7	25.6	**0.0154**	0.007	45.9	ug/L	68	Standard
	Cd	111	43.3	79.3	**0.0106**	0.015	137.7	mg/L	2	Standard
	Cd	114	160.8	66.0	**0.0260**	0.018	70.8	ug/L	12	Standard
>	**In**	115	653140.9	0.6				ug/L	587441	Standard
	Sn	118	840.0	19.7	**0.0347**	0.026	74.7	ug/L	497	Standard
	Sb	123	1054.2	24.1	**0.1991**	0.050	25.2	ug/L	78	Standard
	Ba	135	82.0	54.8	**0.0199**	0.020	97.9	ug/L	24	Standard
	Ce	140	78.3	62.0				ug/L	25	Standard
>	**Tb**	159	1021325.7	0.5				ug/L	939701	Standard
	Ho	165	20.0	66.1				ug/L	10	Standard
	Tl	203	163.0	50.8	**0.0163**	0.009	56.1	ug/L	4	Standard
	Tl	205	543.4	156.5	**0.0696**	0.109	156.0	ug/L	2	Standard
	Pb	206	336.7	16.7	**-0.0136**	0.010	70.5	ug/L	240	Standard
	Pb	207	309.7	12.8	**-0.0104**	0.008	73.2	ug/L	198	Standard
	Pb	208	2162.1	83.2	**0.0313**	0.088	280.1	ug/L	883	Standard
	U	238	524.4	153.4	**0.0652**	0.107	163.8	ug/L	3	Standard
>	**Bi**	209	573163.8	0.1				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:02:12
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.8448**	30.899	173.2	mg/L	0	Standard
Mg	24	123.3	127.6	**0.0829**	0.188	227.1	mg/L	37	Standard
K	39	30.0	66.7	**-0.0524**	0.041	77.8	mg/L	38	Standard
Ca	43	21.7	35.3	**-1.2409**	0.467	37.7	mg/L	10	Standard
Fe	54	185.6	27.6	**0.0086**	0.039	460.9	mg/L	179	Standard
Fe	57	268.3	10.3	**-0.2224**	0.079	35.7	mg/L	335	Standard
Sc-1	45	65305.4	1.2				mg/L	61029	Standard
Cl	35	28806.0	1.7				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1900.1	3.3				ug/L	1547	Standard
P	31	20240.2	2.7				ug/L	15286	Standard
S	34	2435.2	7.1				ug/L	1572	Standard
Sr	88	65.0	20.4				ug/L	73	Standard
C	12	60.0					mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	25.5	84.0				mg/L	6	Standard
Ho-1	165	20.0	66.1				mg/L	10	Standard
Er	166	23.3	65.5				mg/L	13	Standard
I	127	1926.8	28.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		111.783	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:02:12
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.184
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.386
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Mn	55	

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:02:12
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Wednesday, June 29, 2016 18:03:08
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84590.6	2.3				ug/L	74699	Standard
	Be	9	260.0	12.0	**0.2687**	0.034	12.7	ug/L	8	Standard
	Al	27	3603.8	43.5	**0.0082**	0.009	110.1	ug/L	1493	Standard
	Sc	45	65034.3	0.9				ug/L	61029	Standard
	Ti	47	43.7	32.9	**-0.0008**	0.035	4199.9	ug/L	35	Standard
	V	51	4694.6	2.2	**0.3894**	0.008	2.0	ug/L	1435	Standard
	Cr	52	14076.2	0.6	**0.7639**	0.008	1.0	ug/L	7635	Standard
	Cr	53	1435.1	6.7	**1.1315**	0.098	8.6	ug/L	363	Standard
	Mn	55	4379.0	3.5	**0.4438**	0.017	3.9	ug/L	747	Standard
	Co	59	3208.0	4.2	**0.4198**	0.015	3.5	ug/L	234	Standard
	Ni	60	3464.1	1.9	**1.6175**	0.036	2.2	ug/L	215	Standard
	Cu	65	1802.1	7.0	**0.8073**	0.055	6.8	ug/L	104	Standard
	Zn	66	8137.9	1.9	**6.5566**	0.074	1.1	ug/L	363	Standard
>	**Ge**	72	526101.8	0.9				ug/L	495142	Standard
	As	75	520.8	10.5	**0.4970**	0.042	8.4	ug/L	-50	Standard
	Se	82	68.5	3.0	**0.4490**	0.013	3.0	ug/L	14	Standard
	Se-1	77	97.0	14.4	**0.6266**	0.169	26.9	ug/L	42	Standard
>	**Ga**	71	35.0	42.9				mg/L	42	Standard
	Rb	85	53.3	5.4				ug/L	17	Standard
	Y	89	462050.3	3.1				ug/L	417098	Standard
>	**Rh**	103	11.7	24.7				ug/L	13	Standard
	Mo	98	178.6	39.2	**0.0548**	0.024	44.0	ug/L	11	Standard
	Ag	107	2989.6	2.3	**0.4006**	0.011	2.7	ug/L	68	Standard
	Cd	111	543.3	9.6	**0.2264**	0.024	10.6	mg/L	2	Standard
	Cd	114	1493.6	11.4	**0.2605**	0.031	12.1	ug/L	12	Standard
>	**In**	115	651766.1	0.6				ug/L	587441	Standard
	Sn	118	551.7	10.5	**-0.0096**	0.009	97.6	ug/L	497	Standard
	Sb	123	2328.3	1.4	**0.4453**	0.009	1.9	ug/L	78	Standard
	Ba	135	1600.4	3.6	**0.6889**	0.030	4.3	ug/L	24	Standard
	Ce	140	6.7	86.6				ug/L	25	Standard
>	**Tb**	159	1010895.5	0.8				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	860.7	17.5	**0.0936**	0.016	16.7	ug/L	4	Standard
	Tl	205	718.4	10.6	**0.0918**	0.008	8.5	ug/L	2	Standard
	Pb	206	1533.1	9.5	**0.1909**	0.020	10.6	ug/L	240	Standard
	Pb	207	1247.1	5.1	**0.1700**	0.007	4.3	ug/L	198	Standard
	Pb	208	5357.0	3.3	**0.1870**	0.005	2.6	ug/L	883	Standard
	U	238	3040.0	2.0	**0.3994**	0.014	3.4	ug/L	3	Standard
>	**Bi**	209	573346.3	2.5				ug/L	523984	Standard

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 18:05:25
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.2178**	31.546	173.2	mg/L	0	Standard
Mg	24	50.0	50.0	**-0.0044**	0.031	706.6	mg/L	37	Standard
K	39	35.0	24.7	**-0.0415**	0.018	44.0	mg/L	38	Standard
Ca	43	26.7	43.3	**-0.9446**	0.676	71.5	mg/L	10	Standard
Fe	54	105.0	32.5	**-0.0546**	0.028	51.3	mg/L	179	Standard
Fe	57	230.0	24.7	**-0.3395**	0.189	55.5	mg/L	335	Standard
> **Sc-1**	45	65034.3	0.9				mg/L	61029	Standard
Cl	35	27747.4	0.7				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	1776.8	9.7				ug/L	1547	Standard
P	31	14345.1	1.0				ug/L	15286	Standard
S	34	2316.8	2.0				ug/L	1572	Standard
Sr	88	78.3	41.5				ug/L	73	Standard
C	12	70.0	28.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	6.0	100.4				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	1788.4	30.5				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	134.333		
	Al	27			
	Sc	45			
	Ti	47			
	V	51	97.347		
	Cr	52	95.485		
	Cr	53			
	Mn	55	88.762		
	Co	59	104.961		
	Ni	60	101.093		
	Cu	65	100.908		
	Zn	66	104.906		
>	Ge	72		106.253	
	As	75	124.241		
	Se	82	112.254		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 18:05:25
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	100.142		
Cd	111	94.325		
Cd	114			
> In	115			110.950
Sn	118			
Sb	123	111.327		
Ba	135	91.848		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	116.967		
Tl	205			
Pb	206			
Pb	207			
Pb	208	93.500		
U	238	99.844		
> Bi	209			109.420
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 8	Be	9	

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 18:05:25
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBS M9 WG574192-02
Sample Date/Time: Wednesday, June 29, 2016 18:06:21
Number of Replicates: 3
Autosampler Position: 237
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84464.8	1.8				ug/L	74699	Standard
	Be	9	13.3	94.4	**0.0212**	0.013	60.4	ug/L	8	Standard
	Al	27	11484.4	1.6	**0.0557**	0.001	1.2	ug/L	1493	Standard
	Sc	45	65588.4	2.6				ug/L	61029	Standard
	Ti	47	51.3	20.8	**0.0150**	0.026	174.5	ug/L	35	Standard
	V	51	1499.5	1.3	**-0.0137**	0.003	18.5	ug/L	1435	Standard
	Cr	52	8909.3	1.2	**0.0773**	0.005	6.6	ug/L	7635	Standard
	Cr	53	853.4	5.0	**0.5155**	0.034	6.6	ug/L	363	Standard
	Mn	55	1324.4	4.8	**0.0262**	0.011	41.6	ug/L	747	Standard
	Co	59	237.0	9.4	**-0.0044**	0.004	80.7	ug/L	234	Standard
	Ni	60	317.3	8.0	**0.0315**	0.014	46.0	ug/L	215	Standard
	Cu	65	277.0	2.0	**0.0327**	0.005	14.4	ug/L	104	Standard
	Zn	66	2117.5	3.0	**1.3211**	0.055	4.2	ug/L	363	Standard
>	**Ge**	72	540634.9	1.6				ug/L	495142	Standard
	As	75	-19.3	191.5	**0.0441**	0.030	68.3	ug/L	-50	Standard
	Se	82	22.9	30.4	**0.0532**	0.060	113.4	ug/L	14	Standard
	Se-1	77	66.3	3.1	**0.2096**	0.028	13.1	ug/L	42	Standard
>	**Ga**	71	40.0	45.1				mg/L	42	Standard
	Rb	85	45.0	40.1				ug/L	17	Standard
	Y	89	495113.6	1.2				ug/L	417098	Standard
>	**Rh**	103	20.0	50.0				ug/L	13	Standard
	Mo	98	103.4	4.5	**0.0285**	0.002	7.3	ug/L	11	Standard
	Ag	107	105.3	22.1	**0.0013**	0.003	227.8	ug/L	68	Standard
	Cd	111	14.8	35.7	**-0.0017**	0.002	127.2	mg/L	2	Standard
	Cd	114	30.3	51.7	**0.0031**	0.003	86.4	ug/L	12	Standard
>	**In**	115	666604.9	1.7				ug/L	587441	Standard
	Sn	118	860.0	2.1	**0.0351**	0.004	11.7	ug/L	497	Standard
	Sb	123	400.3	27.4	**0.0716**	0.022	30.7	ug/L	78	Standard
	Ba	135	277.0	2.3	**0.1032**	0.001	0.8	ug/L	24	Standard
	Ce	140	75.0	20.0				ug/L	25	Standard
>	**Tb**	159	1002064.8	2.5				ug/L	939701	Standard
	Ho	165	13.3	21.7				ug/L	10	Standard
	Tl	203	23.7	14.8	**0.0009**	0.000	40.1	ug/L	4	Standard
	Tl	205	18.3	56.8	**0.0026**	0.001	51.0	ug/L	2	Standard
	Pb	206	383.3	7.2	**-0.0041**	0.004	93.5	ug/L	240	Standard
	Pb	207	304.7	4.5	**-0.0100**	0.002	23.2	ug/L	198	Standard
	Pb	208	1385.4	1.0	**-0.0050**	0.001	21.8	ug/L	883	Standard
	U	238	16.0	18.8	**-0.0023**	0.000	16.8	ug/L	3	Standard
>	**Bi**	209	560531.0	1.7				ug/L	523984	Standard

Sample ID: PBS M9 WG574192-02
Report Date/Time: Wednesday, June 29, 2016 18:08:38
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	51.7	11.2	**-0.0029**	0.008	286.8	mg/L	37	Standard
K	39	28.3	27.0	**-0.0560**	0.015	26.7	mg/L	38	Standard
Ca	43	25.0	34.6	**-1.0611**	0.483	45.5	mg/L	10	Standard
Fe	54	209.4	8.6	**0.0268**	0.012	44.3	mg/L	179	Standard
Fe	57	206.7	9.2	**-0.4211**	0.048	11.4	mg/L	335	Standard
> **Sc-1**	45	65588.4	2.6				mg/L	61029	Standard
Cl	35	29295.6	1.0				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	3250.3	3.7				ug/L	1547	Standard
P	31	19926.4	1.7				ug/L	15286	Standard
S	34	2431.9	7.4				ug/L	1572	Standard
Sr	88	63.3	32.9				ug/L	73	Standard
C	12	60.0	16.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	23.2	64.8				mg/L	6	Standard
Ho-1	165	13.3	21.7				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	1665.1	25.5				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		113.074	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		109.188	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBS M9 WG574192-02
Report Date/Time: Wednesday, June 29, 2016 18:08:38
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.476
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.975
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBS M9 WG574192-02
Report Date/Time: Wednesday, June 29, 2016 18:08:38
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSS M9 WG574192-03
Sample Date/Time: Wednesday, June 29, 2016 18:09:32
Number of Replicates: 3
Autosampler Position: 238
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83350.4	2.2				ug/L	74699	Standard
	Be	9	28323.1	1.8	**28.8492**	0.865	3.0	ug/L	8	Standard
	Al	27	10196.8	1.6	**0.0488**	0.002	4.6	ug/L	1493	Standard
	Sc	45	65183.2	0.6				ug/L	61029	Standard
	Ti	47	95.7	13.6	**0.1113**	0.027	24.3	ug/L	35	Standard
	V	51	224256.1	0.6	**25.7588**	0.522	2.0	ug/L	1435	Standard
	Cr	52	224696.6	1.0	**25.4792**	0.636	2.5	ug/L	7635	Standard
	Cr	53	27321.3	1.7	**25.5112**	0.743	2.9	ug/L	363	Standard
	Mn	55	202745.7	1.5	**25.3082**	0.750	3.0	ug/L	747	Standard
	Co	59	203564.8	2.3	**26.9126**	1.011	3.8	ug/L	234	Standard
	Ni	60	54379.7	0.7	**25.2634**	0.489	1.9	ug/L	215	Standard
	Cu	65	53975.6	1.0	**25.2465**	0.617	2.4	ug/L	104	Standard
	Zn	66	33582.8	1.1	**26.4199**	0.666	2.5	ug/L	363	Standard
>	**Ge**	72	566570.7	1.5				ug/L	495142	Standard
	As	75	31915.6	1.0	**24.9643**	0.450	1.8	ug/L	-50	Standard
	Se	82	3210.1	2.6	**25.4315**	0.960	3.8	ug/L	14	Standard
	Se-1	77	2232.8	4.1	**26.1060**	1.397	5.4	ug/L	42	Standard
>	**Ga**	71	91.7	13.7				mg/L	42	Standard
	Rb	85	96.7	31.2				ug/L	17	Standard
	Y	89	463441.3	2.1				ug/L	417098	Standard
>	**Rh**	103	25.0	40.0				ug/L	13	Standard
	Mo	98	76.3	5.4	**0.0207**	0.001	6.1	ug/L	11	Standard
	Ag	107	186996.6	1.0	**26.4732**	0.339	1.3	ug/L	68	Standard
	Cd	111	58916.9	1.6	**26.0048**	0.541	2.1	mg/L	2	Standard
	Cd	114	141808.0	1.8	**25.5238**	0.616	2.4	ug/L	12	Standard
>	**In**	115	636399.6	0.6				ug/L	587441	Standard
	Sn	118	665.0	9.8	**0.0103**	0.010	94.1	ug/L	497	Standard
	Sb	123	127977.1	1.2	**25.3000**	0.453	1.8	ug/L	78	Standard
	Ba	135	55490.8	0.9	**25.0111**	0.350	1.4	ug/L	24	Standard
	Ce	140	61.7	36.6				ug/L	25	Standard
>	**Tb**	159	977939.2	1.1				ug/L	939701	Standard
	Ho	165	5.0	100.0				ug/L	10	Standard
	Tl	203	232757.9	0.8	**26.2994**	0.658	2.5	ug/L	4	Standard
	Tl	205	213539.2	0.1	**27.8001**	0.582	2.1	ug/L	2	Standard
	Pb	206	150898.4	0.7	**26.2407**	0.649	2.5	ug/L	240	Standard
	Pb	207	129115.2	1.2	**25.2720**	0.800	3.2	ug/L	198	Standard
	Pb	208	547777.7	0.6	**27.1359**	0.615	2.3	ug/L	883	Standard
	U	238	197609.3	0.3	**26.7341**	0.481	1.8	ug/L	3	Standard
>	**Bi**	209	562718.9	2.0				ug/L	523984	Standard

Sample ID: LCSS M9 WG574192-03
Report Date/Time: Wednesday, June 29, 2016 18:11:49
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**35.9815**	31.158	86.6	mg/L	0	Standard
Mg	24	53.3	23.6	**-0.0007**	0.015	2138.9	mg/L	37	Standard
K	39	36.7	90.8	**-0.0380**	0.070	183.0	mg/L	38	Standard
Ca	43	26.7	28.6	**-0.9444**	0.462	48.9	mg/L	10	Standard
Fe	54	149.2	16.7	**-0.0198**	0.020	102.0	mg/L	179	Standard
Fe	57	258.3	10.7	**-0.2518**	0.092	36.6	mg/L	335	Standard
> **Sc-1**	45	65183.2	0.6				mg/L	61029	Standard
Cl	35	28068.6	2.1				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	1853.4	3.5				ug/L	1547	Standard
P	31	20221.8	1.7				ug/L	15286	Standard
S	34	2256.8	4.8				ug/L	1572	Standard
Sr	88	75.0	24.0				ug/L	73	Standard
C	12	113.3	53.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	2.9	203.0				mg/L	6	Standard
Ho-1	165	5.0	100.0				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	1621.8	26.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.582	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.426	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSS M9 WG574192-03
Report Date/Time: Wednesday, June 29, 2016 18:11:49
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.334
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.392
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSS M9 WG574192-03
Report Date/Time: Wednesday, June 29, 2016 18:11:49
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120601 WG574192-01
Sample Date/Time: Wednesday, June 29, 2016 18:12:44
Number of Replicates: 3
Autosampler Position: 239
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89463.4	1.2				ug/L	74699	Standard
	Be	9	386.7	13.6	**0.3740**	0.046	12.2	ug/L	8	Standard
	Al	27	7605562.2	1.6	**43.1976**	0.340	0.8	ug/L	1493	Standard
	Sc	45	72243.5	1.6				ug/L	61029	Standard
	Ti	47	9362.6	0.8	**21.8169**	0.679	3.1	ug/L	35	Standard
	V	51	123334.5	1.4	**14.2946**	0.165	1.2	ug/L	1435	Standard
	Cr	52	107034.1	1.6	**11.7966**	0.478	4.1	ug/L	7635	Standard
	Cr	53	12683.6	2.9	**11.8556**	0.642	5.4	ug/L	363	Standard
	Mn	55	2726083.5	2.1	**347.2409**	2.026	0.6	ug/L	747	Standard
	Co	59	29531.4	3.0	**3.9299**	0.043	1.1	ug/L	234	Standard
	Ni	60	23196.4	0.4	**10.8725**	0.312	2.9	ug/L	215	Standard
	Cu	65	77156.3	0.7	**36.6848**	0.692	1.9	ug/L	104	Standard
	Zn	66	748003.6	1.6	**607.1548**	7.539	1.2	ug/L	363	Standard
>	**Ge**	72	558122.4	2.5				ug/L	495142	Standard
	As	75	3312.8	2.4	**2.6835**	0.016	0.6	ug/L	-50	Standard
	Se	82	112.5	6.7	**0.7726**	0.083	10.8	ug/L	14	Standard
	Se-1	77	130.7	4.7	**0.9656**	0.087	9.0	ug/L	42	Standard
>	**Ga**	71	7793.7	3.0				mg/L	42	Standard
	Rb	85	97459.1	1.1				ug/L	17	Standard
	Y	89	585281.5	2.1				ug/L	417098	Standard
>	**Rh**	103	56.7	13.5				ug/L	13	Standard
	Mo	98	1826.8	4.5	**0.5887**	0.034	5.8	ug/L	11	Standard
	Ag	107	1189.0	7.2	**0.1453**	0.012	8.4	ug/L	68	Standard
	Cd	111	5739.0	1.3	**2.3714**	0.004	0.2	mg/L	2	Standard
	Cd	114	13674.5	3.2	**2.3090**	0.056	2.4	ug/L	12	Standard
>	**In**	115	677670.9	1.2				ug/L	587441	Standard
	Sn	118	2238.5	6.9	**0.2377**	0.027	11.3	ug/L	497	Standard
	Sb	123	1060.8	22.9	**0.1924**	0.043	22.3	ug/L	78	Standard
	Ba	135	275486.3	0.7	**116.6679**	1.279	1.1	ug/L	24	Standard
	Ce	140	521396.3	0.5				ug/L	25	Standard
>	**Tb**	159	1067665.8	1.0				ug/L	939701	Standard
	Ho	165	8012.1	4.7				ug/L	10	Standard
	Tl	203	658.7	5.9	**0.0698**	0.004	5.8	ug/L	4	Standard
	Tl	205	588.3	7.1	**0.0738**	0.005	7.2	ug/L	2	Standard
	Pb	206	280687.1	0.1	**46.9821**	0.294	0.6	ug/L	240	Standard
	Pb	207	217657.6	0.6	**40.9978**	0.453	1.1	ug/L	198	Standard
	Pb	208	937818.9	1.0	**44.7144**	0.652	1.5	ug/L	883	Standard
	U	238	5450.3	0.5	**0.7046**	0.003	0.4	ug/L	3	Standard
>	**Bi**	209	585148.7	0.6				ug/L	523984	Standard

Sample ID: L1606120601 WG574192-01
Report Date/Time: Wednesday, June 29, 2016 18:15:01
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	733.4	5.7	**0.7343**	0.035	4.8	mg/L	37	Standard
K	39	401.7	1.9	**0.6360**	0.019	3.0	mg/L	38	Standard
Ca	43	643.3	17.5	**31.7962**	5.845	18.4	mg/L	10	Standard
Fe	54	8818.1	2.4	**6.1777**	0.130	2.1	mg/L	179	Standard
Fe	57	2673.6	4.7	**6.5999**	0.243	3.7	mg/L	335	Standard
> **Sc-1**	45	72243.5	1.6				mg/L	61029	Standard
Cl	35	29798.6	1.4				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	3707.1	5.8				ug/L	1547	Standard
P	31	20168.4	2.0				ug/L	15286	Standard
S	34	2373.5	6.9				ug/L	1572	Standard
Sr	88	256.7	13.8				ug/L	73	Standard
C	12	350.0	7.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	130.0	53.3				mg/L	3	Standard
Dy	164	13278.3	0.5				mg/L	6	Standard
Ho-1	165	8012.1	4.7				mg/L	10	Standard
Er	166	7608.6	3.8				mg/L	13	Standard
I	127	89948.0	2.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		119.765	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		112.720	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120601 WG574192-01
Report Date/Time: Wednesday, June 29, 2016 18:15:01
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		115.360
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.673
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120601 WG574192-01
Report Date/Time: Wednesday, June 29, 2016 18:15:01
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120602
Sample Date/Time: Wednesday, June 29, 2016 18:15:55
Number of Replicates: 3
Autosampler Position: 240
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88784.4	1.1				ug/L	74699	Standard
	Be	9	281.7	5.7	**0.2767**	0.013	4.7	ug/L	8	Standard
	Al	27	7005404.7	2.4	**40.0898**	0.625	1.6	ug/L	1493	Standard
	Sc	45	70594.2	3.3				ug/L	61029	Standard
	Ti	47	10155.7	1.6	**24.3492**	0.468	1.9	ug/L	35	Standard
	V	51	153110.0	0.2	**18.3086**	0.101	0.6	ug/L	1435	Standard
	Cr	52	102957.6	0.7	**11.6568**	0.092	0.8	ug/L	7635	Standard
	Cr	53	11732.9	3.0	**11.2567**	0.305	2.7	ug/L	363	Standard
	Mn	55	2052735.8	1.1	**268.9604**	2.146	0.8	ug/L	747	Standard
	Co	59	33492.3	1.4	**4.5921**	0.057	1.2	ug/L	234	Standard
	Ni	60	26642.1	0.7	**12.8642**	0.053	0.4	ug/L	215	Standard
	Cu	65	40463.3	0.7	**19.7403**	0.183	0.9	ug/L	104	Standard
	Zn	66	608360.3	0.1	**507.8962**	1.599	0.3	ug/L	363	Standard
>	**Ge**	72	542468.6	0.4				ug/L	495142	Standard
	As	75	4231.0	2.7	**3.5076**	0.102	2.9	ug/L	-50	Standard
	Se	82	256.2	4.5	**1.9932**	0.099	5.0	ug/L	14	Standard
	Se-1	77	208.7	7.7	**1.9860**	0.210	10.6	ug/L	42	Standard
>	**Ga**	71	5872.8	1.9				mg/L	42	Standard
	Rb	85	68412.5	1.5				ug/L	17	Standard
	Y	89	570393.2	0.3				ug/L	417098	Standard
>	**Rh**	103	206.7	25.4				ug/L	13	Standard
	Mo	98	11374.8	0.4	**3.8637**	0.075	1.9	ug/L	11	Standard
	Ag	107	1353.7	2.3	**0.1754**	0.001	0.8	ug/L	68	Standard
	Cd	111	5491.2	1.2	**2.3732**	0.046	1.9	mg/L	2	Standard
	Cd	114	13400.9	1.9	**2.3668**	0.045	1.9	ug/L	12	Standard
>	**In**	115	648040.3	1.6				ug/L	587441	Standard
	Sn	118	1990.1	9.0	**0.2145**	0.033	15.4	ug/L	497	Standard
	Sb	123	589.0	13.2	**0.1100**	0.013	12.1	ug/L	78	Standard
	Ba	135	226861.9	1.0	**100.4795**	2.067	2.1	ug/L	24	Standard
	Ce	140	495305.9	0.7				ug/L	25	Standard
>	**Tb**	159	1028084.3	0.7				ug/L	939701	Standard
	Ho	165	7586.9	1.7				ug/L	10	Standard
	Tl	203	574.7	1.8	**0.0640**	0.002	2.9	ug/L	4	Standard
	Tl	205	501.7	7.3	**0.0663**	0.004	6.1	ug/L	2	Standard
	Pb	206	2490204.6	1.0	**439.6443**	9.561	2.2	ug/L	240	Standard
	Pb	207	1941338.7	1.0	**385.7578**	8.533	2.2	ug/L	198	Standard
	Pb	208	8788859.7	0.3	**442.0013**	4.418	1.0	ug/L	883	Standard
	U	238	27264.2	0.8	**3.7315**	0.055	1.5	ug/L	3	Standard
>	**Bi**	209	555601.8	1.3				ug/L	523984	Standard

Sample ID: L1606120602
Report Date/Time: Wednesday, June 29, 2016 18:18:12
Page 1

Approved: July 04, 2016

Na	23	15.0	66.7	**151.5495**	103.438	68.3	mg/L	0	Standard
Mg	24	875.0	7.4	**0.9112**	0.062	6.8	mg/L	37	Standard
K	39	308.3	17.3	**0.4774**	0.118	24.7	mg/L	38	Standard
Ca	43	2528.5	1.9	**135.7080**	7.063	5.2	mg/L	10	Standard
Fe	54	8423.7	4.4	**6.0457**	0.457	7.6	mg/L	179	Standard
Fe	57	3312.0	6.3	**8.6682**	0.763	8.8	mg/L	335	Standard
Sc-1	45	70594.2	3.3				mg/L	61029	Standard
Cl	35	33743.5	0.4				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	4283.9	2.2				ug/L	1547	Standard
P	31	21890.8	1.3				ug/L	15286	Standard
S	34	2663.6	6.3				ug/L	1572	Standard
Sr	88	533.3	8.7				ug/L	73	Standard
C	12	600.0	5.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	100.0	26.5				mg/L	3	Standard
Dy	164	11590.5	1.3				mg/L	6	Standard
Ho-1	165	7586.9	1.7				mg/L	10	Standard
Er	166	6698.1	4.4				mg/L	13	Standard
I	127	108507.7	4.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		118.857	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		109.558	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120602
Report Date/Time: Wednesday, June 29, 2016 18:18:12
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	110.316
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	106.034
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120602
Report Date/Time: Wednesday, June 29, 2016 18:18:12
Page 3

Approved: July 04, 2016

Pb 206 Upper, S, EEE	Pb	206
Pb 207 Upper, S, EEE	Pb	207
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120603S WG574192-04
Sample Date/Time: Wednesday, June 29, 2016 18:19:07
Number of Replicates: 3
Autosampler Position: 241
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86561.6	0.8				ug/L	74699	Standard
	Be	9	28455.1	4.4	**27.8947**	1.031	3.7	ug/L	8	Standard
	Al	27	11326081.2	0.4	**66.4980**	0.812	1.2	ug/L	1493	Standard
	Sc	45	74033.8	1.5				ug/L	61029	Standard
	Ti	47	13830.6	1.3	**31.7218**	0.771	2.4	ug/L	35	Standard
	V	51	404083.1	0.7	**46.4723**	0.823	1.8	ug/L	1435	Standard
	Cr	52	396378.6	0.5	**45.6257**	0.448	1.0	ug/L	7635	Standard
	Cr	53	48917.2	0.9	**45.8335**	0.539	1.2	ug/L	363	Standard
	Mn	55	3610519.8	1.3	**452.1326**	4.314	1.0	ug/L	747	Standard
	Co	59	243070.7	1.0	**32.0670**	0.356	1.1	ug/L	234	Standard
	Ni	60	87954.6	1.3	**40.8489**	0.598	1.5	ug/L	215	Standard
	Cu	65	138890.7	1.2	**64.9805**	0.614	0.9	ug/L	104	Standard
	Zn	66	982893.1	0.9	**784.3980**	13.633	1.7	ug/L	363	Standard
>	**Ge**	72	567734.3	1.2				ug/L	495142	Standard
	As	75	35264.6	1.3	**27.5210**	0.578	2.1	ug/L	-50	Standard
	Se	82	3103.1	1.8	**24.5237**	0.601	2.4	ug/L	14	Standard
	Se-1	77	2151.2	2.8	**25.0663**	0.745	3.0	ug/L	42	Standard
>	**Ga**	71	14508.6	3.6				mg/L	42	Standard
	Rb	85	180073.9	1.5				ug/L	17	Standard
	Y	89	616764.0	1.9				ug/L	417098	Standard
>	**Rh**	103	85.0	11.8				ug/L	13	Standard
	Mo	98	3441.4	2.1	**1.1615**	0.033	2.8	ug/L	11	Standard
	Ag	107	170180.4	0.7	**23.5897**	0.078	0.3	ug/L	68	Standard
	Cd	111	63817.8	1.8	**27.5852**	0.736	2.7	mg/L	2	Standard
	Cd	114	155790.6	2.3	**27.4604**	0.877	3.2	ug/L	12	Standard
>	**In**	115	649912.7	0.9				ug/L	587441	Standard
	Sn	118	2755.3	3.8	**0.3317**	0.018	5.5	ug/L	497	Standard
	Sb	123	11515.0	1.7	**2.2251**	0.028	1.3	ug/L	78	Standard
	Ba	135	420339.6	0.5	**185.6246**	2.179	1.2	ug/L	24	Standard
	Ce	140	771001.7	1.2				ug/L	25	Standard
>	**Tb**	159	1033151.8	0.5				ug/L	939701	Standard
	Ho	165	11344.2	1.5				ug/L	10	Standard
	Tl	203	220759.0	0.5	**24.6088**	0.113	0.5	ug/L	4	Standard
	Tl	205	203485.6	0.4	**26.1371**	0.140	0.5	ug/L	2	Standard
	Pb	206	518067.3	0.6	**89.0511**	0.447	0.5	ug/L	240	Standard
	Pb	207	415032.4	0.7	**80.2896**	0.537	0.7	ug/L	198	Standard
	Pb	208	1813547.0	1.2	**88.8061**	1.200	1.4	ug/L	883	Standard
	U	238	196913.8	1.4	**26.2850**	0.405	1.5	ug/L	3	Standard
>	**Bi**	209	570191.4	0.2				ug/L	523984	Standard

Sample ID: L1606120603S WG574192-04
Report Date/Time: Wednesday, June 29, 2016 18:21:24
Page 1

Approved: July 04, 2016

Na	23	6.7	43.3	**63.0881**	26.284	41.7	mg/L	0	Standard
Mg	24	761.7	8.6	**0.7448**	0.057	7.6	mg/L	37	Standard
K	39	698.3	9.3	**1.1574**	0.103	8.9	mg/L	38	Standard
Ca	43	761.7	2.5	**37.1357**	0.619	1.7	mg/L	10	Standard
Fe	54	13661.1	2.3	**9.4129**	0.352	3.7	mg/L	179	Standard
Fe	57	3945.5	4.9	**9.9839**	0.617	6.2	mg/L	335	Standard
> **Sc-1**	45	74033.8	1.5				mg/L	61029	Standard
Cl	35	29672.4	3.3				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	5120.9	2.8				ug/L	1547	Standard
P	31	21184.8	2.3				ug/L	15286	Standard
S	34	2585.2	3.6				ug/L	1572	Standard
Sr	88	278.3	7.3				ug/L	73	Standard
C	12	453.3	23.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	180.0	28.9				mg/L	3	Standard
Dy	164	18295.9	3.4				mg/L	6	Standard
Ho-1	165	11344.2	1.5				mg/L	10	Standard
Er	166	10350.2	2.2				mg/L	13	Standard
I	127	155575.8	4.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		115.881	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.661	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120603S WG574192-04
Report Date/Time: Wednesday, June 29, 2016 18:21:24
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	110.635
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.818
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120603S WG574192-04
Report Date/Time: Wednesday, June 29, 2016 18:21:24
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120604SD WG574192-05
Sample Date/Time: Wednesday, June 29, 2016 18:22:18
Number of Replicates: 3
Autosampler Position: 242
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88318.4	2.5				ug/L	74699	Standard
	Be	9	28910.9	0.8	**27.7911**	0.552	2.0	ug/L	8	Standard
	Al	27	11678142.6	0.8	**67.2186**	1.350	2.0	ug/L	1493	Standard
	Sc	45	75929.8	1.2				ug/L	61029	Standard
	Ti	47	15202.6	0.9	**34.3685**	0.367	1.1	ug/L	35	Standard
	V	51	408247.6	0.9	**46.2682**	0.632	1.4	ug/L	1435	Standard
	Cr	52	411599.1	1.0	**46.7151**	0.586	1.3	ug/L	7635	Standard
	Cr	53	50226.5	2.4	**46.3783**	0.843	1.8	ug/L	363	Standard
	Mn	55	4064951.2	0.1	**501.7134**	5.151	1.0	ug/L	747	Standard
	Co	59	242783.8	0.7	**31.5662**	0.473	1.5	ug/L	234	Standard
	Ni	60	92537.7	0.8	**42.3583**	0.203	0.5	ug/L	215	Standard
	Cu	65	135365.1	1.4	**62.4070**	0.241	0.4	ug/L	104	Standard
	Zn	66	1025151.4	1.0	**806.2133**	6.015	0.7	ug/L	363	Standard
>	**Ge**	72	576071.4	1.0				ug/L	495142	Standard
	As	75	35892.9	0.1	**27.6040**	0.298	1.1	ug/L	-50	Standard
	Se	82	2975.5	0.8	**23.1647**	0.155	0.7	ug/L	14	Standard
	Se-1	77	2118.1	0.7	**24.3051**	0.149	0.6	ug/L	42	Standard
>	**Ga**	71	16053.5	1.5				mg/L	42	Standard
	Rb	85	192660.5	0.7				ug/L	17	Standard
	Y	89	633758.5	0.9				ug/L	417098	Standard
>	**Rh**	103	91.7	22.7				ug/L	13	Standard
	Mo	98	2104.9	2.7	**0.6926**	0.023	3.3	ug/L	11	Standard
	Ag	107	176068.8	0.4	**23.8725**	0.248	1.0	ug/L	68	Standard
	Cd	111	64701.8	0.7	**27.3512**	0.177	0.6	mg/L	2	Standard
	Cd	114	157399.0	1.1	**27.1310**	0.176	0.6	ug/L	12	Standard
>	**In**	115	664480.0	1.2				ug/L	587441	Standard
	Sn	118	2676.9	2.8	**0.3105**	0.015	4.8	ug/L	497	Standard
	Sb	123	10363.1	2.2	**1.9584**	0.055	2.8	ug/L	78	Standard
	Ba	135	439420.6	0.3	**189.8023**	2.264	1.2	ug/L	24	Standard
	Ce	140	842883.5	1.0				ug/L	25	Standard
>	**Tb**	159	1049787.1	1.3				ug/L	939701	Standard
	Ho	165	12211.6	1.7				ug/L	10	Standard
	Tl	203	222058.4	0.3	**24.4396**	0.316	1.3	ug/L	4	Standard
	Tl	205	202828.8	1.6	**25.7185**	0.095	0.4	ug/L	2	Standard
	Pb	206	574235.7	0.3	**97.4667**	1.839	1.9	ug/L	240	Standard
	Pb	207	457211.7	0.6	**87.3408**	1.861	2.1	ug/L	198	Standard
	Pb	208	1979847.3	0.4	**95.7233**	1.171	1.2	ug/L	883	Standard
	U	238	197274.4	0.9	**26.0018**	0.641	2.5	ug/L	3	Standard
>	**Bi**	209	577596.9	1.6				ug/L	523984	Standard

Sample ID: L1606120604SD WG574192-05
Report Date/Time: Wednesday, June 29, 2016 18:24:35
Page 1

Approved: July 04, 2016

Na	23	5.0	0.0	**46.3267**	0.533	1.2	mg/L	0	Standard
Mg	24	961.7	5.5	**0.9325**	0.047	5.0	mg/L	37	Standard
K	39	738.4	15.4	**1.1991**	0.216	18.0	mg/L	38	Standard
Ca	43	858.4	11.4	**41.0973**	5.443	13.2	mg/L	10	Standard
Fe	54	16143.5	2.3	**10.8618**	0.122	1.1	mg/L	179	Standard
Fe	57	4519.0	2.4	**11.2744**	0.427	3.8	mg/L	335	Standard
> **Sc-1**	45	75929.8	1.2				mg/L	61029	Standard
Cl	35	29136.0	1.4				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	5384.3	8.1				ug/L	1547	Standard
P	31	20994.5	3.6				ug/L	15286	Standard
S	34	2526.9	1.1				ug/L	1572	Standard
Sr	88	265.0	8.2				ug/L	73	Standard
C	12	466.7	15.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	193.3	16.6				mg/L	3	Standard
Dy	164	19678.1	2.3				mg/L	6	Standard
Ho-1	165	12211.6	1.7				mg/L	10	Standard
Er	166	11454.3	2.8				mg/L	13	Standard
I	127	133210.9	5.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		118.233	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		116.345	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120604SD WG574192-05
Report Date/Time: Wednesday, June 29, 2016 18:24:35
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	113.114
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	110.232
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120604SD WG574192-05
Report Date/Time: Wednesday, June 29, 2016 18:24:35
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605
Sample Date/Time: Wednesday, June 29, 2016 18:25:30
Number of Replicates: 3
Autosampler Position: 243
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90694.1	2.2				ug/L	74699	Standard
	Be	9	526.7	6.5	**0.5004**	0.031	6.1	ug/L	8	Standard
	Al	27	8237088.6	1.8	**46.1607**	0.953	2.1	ug/L	1493	Standard
	Sc	45	76241.4	2.3				ug/L	61029	Standard
	Ti	47	9939.3	1.0	**23.0163**	0.554	2.4	ug/L	35	Standard
	V	51	146708.3	2.5	**16.9369**	0.672	4.0	ug/L	1435	Standard
	Cr	52	214286.2	1.5	**24.4749**	0.700	2.9	ug/L	7635	Standard
	Cr	53	25651.7	1.8	**24.1415**	0.427	1.8	ug/L	363	Standard
	Mn	55	2369893.5	2.4	**300.0787**	12.076	4.0	ug/L	747	Standard
	Co	59	42091.1	2.2	**5.5844**	0.210	3.8	ug/L	234	Standard
	Ni	60	29674.7	1.4	**13.8532**	0.382	2.8	ug/L	215	Standard
	Cu	65	80108.6	1.5	**37.8555**	1.188	3.1	ug/L	104	Standard
	Zn	66	402296.4	1.3	**324.3434**	9.485	2.9	ug/L	363	Standard
>	**Ge**	72	561626.1	1.7				ug/L	495142	Standard
	As	75	4613.8	0.6	**3.6915**	0.048	1.3	ug/L	-50	Standard
	Se	82	73.3	11.4	**0.4501**	0.057	12.7	ug/L	14	Standard
	Se-1	77	107.3	9.4	**0.6736**	0.125	18.5	ug/L	42	Standard
>	**Ga**	71	10595.4	4.5				mg/L	42	Standard
	Rb	85	117987.7	1.5				ug/L	17	Standard
	Y	89	650208.5	0.3				ug/L	417098	Standard
>	**Rh**	103	85.0	38.6				ug/L	13	Standard
	Mo	98	1074.4	2.1	**0.3404**	0.008	2.3	ug/L	11	Standard
	Ag	107	1521.7	2.8	**0.1876**	0.005	2.8	ug/L	68	Standard
	Cd	111	14943.9	2.9	**6.1278**	0.124	2.0	mg/L	2	Standard
	Cd	114	36335.4	1.3	**6.0807**	0.042	0.7	ug/L	12	Standard
>	**In**	115	684218.1	1.0				ug/L	587441	Standard
	Sn	118	4600.7	0.6	**0.5815**	0.007	1.3	ug/L	497	Standard
	Sb	123	346.8	6.2	**0.0596**	0.003	5.7	ug/L	78	Standard
	Ba	135	305952.3	0.8	**128.3272**	1.108	0.9	ug/L	24	Standard
	Ce	140	926252.0	0.8				ug/L	25	Standard
>	**Tb**	159	1067500.9	0.3				ug/L	939701	Standard
	Ho	165	12695.3	1.4				ug/L	10	Standard
	Tl	203	780.0	9.1	**0.0850**	0.009	10.5	ug/L	4	Standard
	Tl	205	700.0	0.7	**0.0898**	0.001	1.1	ug/L	2	Standard
	Pb	206	317389.9	0.7	**54.3657**	1.082	2.0	ug/L	240	Standard
	Pb	207	249270.6	1.1	**48.0507**	1.100	2.3	ug/L	198	Standard
	Pb	208	1104553.9	2.1	**53.8955**	1.488	2.8	ug/L	883	Standard
	U	238	5524.0	2.1	**0.7309**	0.022	3.0	ug/L	3	Standard
>	**Bi**	209	572007.3	1.4				ug/L	523984	Standard

Sample ID: L1606120605
Report Date/Time: Wednesday, June 29, 2016 18:27:47
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**77.0658**	27.245	35.4	mg/L	0	Standard
Mg	24	545.0	4.9	**0.4986**	0.038	7.7	mg/L	37	Standard
K	39	528.3	15.8	**0.8195**	0.134	16.4	mg/L	38	Standard
Ca	43	966.7	6.0	**46.3930**	3.592	7.7	mg/L	10	Standard
Fe	54	12851.8	4.3	**8.5861**	0.399	4.6	mg/L	179	Standard
Fe	57	3822.2	7.3	**9.3402**	1.007	10.8	mg/L	335	Standard
Sc-1	45	76241.4	2.3				mg/L	61029	Standard
Cl	35	29717.8	2.2				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	3207.0	9.1				ug/L	1547	Standard
P	31	20467.1	0.8				ug/L	15286	Standard
S	34	2505.2	3.0				ug/L	1572	Standard
Sr	88	263.3	8.6				ug/L	73	Standard
C	12	196.7	5.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	123.3	46.8				mg/L	3	Standard
Dy	164	19864.2	0.8				mg/L	6	Standard
Ho-1	165	12695.3	1.4				mg/L	10	Standard
Er	166	11821.3	4.6				mg/L	13	Standard
I	127	41200.9	0.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		121.413	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.427	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120605
Report Date/Time: Wednesday, June 29, 2016 18:27:47
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.474
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.165
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606120605
Report Date/Time: Wednesday, June 29, 2016 18:27:47
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605PS WG574690-01
Sample Date/Time: Wednesday, June 29, 2016 18:28:41
Number of Replicates: 3
Autosampler Position: 244
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85371.5	1.3				ug/L	74699	Standard
	Be	9	57838.5	0.7	**57.4989**	0.995	1.7	ug/L	8	Standard
	Al	27	8190235.3	0.9	**48.7593**	1.052	2.2	ug/L	1493	Standard
	Sc	45	74174.5	1.1				ug/L	61029	Standard
	Ti	47	9905.9	1.1	**22.7993**	0.140	0.6	ug/L	35	Standard
	V	51	615589.0	1.1	**71.2575**	1.604	2.3	ug/L	1435	Standard
	Cr	52	662720.1	2.0	**77.3523**	0.990	1.3	ug/L	7635	Standard
	Cr	53	81212.1	1.6	**76.6855**	0.228	0.3	ug/L	363	Standard
	Mn	55	2850355.8	1.2	**358.7372**	7.553	2.1	ug/L	747	Standard
	Co	59	453959.9	1.8	**60.2304**	1.747	2.9	ug/L	234	Standard
	Ni	60	139316.4	1.9	**65.0924**	0.607	0.9	ug/L	215	Standard
	Cu	65	189057.7	1.5	**88.9267**	0.405	0.5	ug/L	104	Standard
	Zn	66	476114.4	1.0	**381.6097**	5.307	1.4	ug/L	363	Standard
>	**Ge**	72	564919.1	1.3				ug/L	495142	Standard
	As	75	71672.0	0.5	**56.1469**	0.522	0.9	ug/L	-50	Standard
	Se	82	6681.0	0.4	**53.2203**	0.482	0.9	ug/L	14	Standard
	Se-1	77	4575.0	1.8	**54.2855**	1.161	2.1	ug/L	42	Standard
>	**Ga**	71	10925.6	2.3				mg/L	42	Standard
	Rb	85	119128.2	1.4				ug/L	17	Standard
	Y	89	658944.3	2.5				ug/L	417098	Standard
>	**Rh**	103	98.3	24.0				ug/L	13	Standard
	Mo	98	1073.3	2.3	**0.3534**	0.015	4.1	ug/L	11	Standard
	Ag	107	364831.9	0.7	**49.9049**	1.416	2.8	ug/L	68	Standard
	Cd	111	132655.1	1.4	**56.5569**	1.029	1.8	mg/L	2	Standard
	Cd	114	321468.3	0.7	**55.8946**	1.680	3.0	ug/L	12	Standard
>	**In**	115	659074.5	2.5				ug/L	587441	Standard
	Sn	118	1318.4	1.7	**0.1066**	0.008	7.5	ug/L	497	Standard
	Sb	123	258576.1	0.4	**49.3799**	1.200	2.4	ug/L	78	Standard
	Ba	135	409891.1	0.5	**178.5483**	4.068	2.3	ug/L	24	Standard
	Ce	140	929186.0	0.4				ug/L	25	Standard
>	**Tb**	159	1023915.1	1.8				ug/L	939701	Standard
	Ho	165	12935.5	3.1				ug/L	10	Standard
	Tl	203	455240.8	0.8	**52.4471**	0.837	1.6	ug/L	4	Standard
	Tl	205	420642.4	0.2	**55.8346**	0.392	0.7	ug/L	2	Standard
	Pb	206	614831.5	0.9	**109.2366**	1.843	1.7	ug/L	240	Standard
	Pb	207	503849.3	0.9	**100.7500**	1.617	1.6	ug/L	198	Standard
	Pb	208	2170185.0	1.2	**109.8441**	2.227	2.0	ug/L	883	Standard
	U	238	393858.5	1.0	**54.3381**	1.023	1.9	ug/L	3	Standard
>	**Bi**	209	551780.7	0.8				ug/L	523984	Standard

Sample ID: L1606120605PS WG574690-01
Report Date/Time: Wednesday, June 29, 2016 18:30:58
Page 1

Approved: July 04, 2016

Na	23	10.0	86.6	**95.2332**	83.153	87.3	mg/L	0	Standard
Mg	24	453.3	5.6	**0.4164**	0.027	6.4	mg/L	37	Standard
K	39	480.0	7.5	**0.7589**	0.067	8.8	mg/L	38	Standard
Ca	43	940.0	10.6	**46.3668**	5.650	12.2	mg/L	10	Standard
Fe	54	12349.5	1.2	**8.4763**	0.023	0.3	mg/L	179	Standard
Fe	57	3702.1	2.7	**9.2777**	0.167	1.8	mg/L	335	Standard
Sc-1	45	74174.5	1.1				mg/L	61029	Standard
Cl	35	30087.8	0.9				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	3260.4	2.1				ug/L	1547	Standard
P	31	20787.6	1.7				ug/L	15286	Standard
S	34	2500.2	2.8				ug/L	1572	Standard
Sr	88	245.0	12.4				ug/L	73	Standard
C	12	160.0	54.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	103.3	22.3				mg/L	3	Standard
Dy	164	20338.3	3.5				mg/L	6	Standard
Ho-1	165	12935.5	3.1				mg/L	10	Standard
Er	166	11817.9	3.9				mg/L	13	Standard
I	127	39775.5	3.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		114.288	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.092	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120605PS WG574690-01
Report Date/Time: Wednesday, June 29, 2016 18:30:58
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	112.194
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	105.305
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120605PS WG574690-01
Report Date/Time: Wednesday, June 29, 2016 18:30:58
Page 3

Approved: July 04, 2016

Pb 206 Upper, S, EEE	Pb	206
Pb 207 Upper, S, EEE	Pb	207
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605SDL WG574690-02
Sample Date/Time: Wednesday, June 29, 2016 18:31:53
Number of Replicates: 3
Autosampler Position: 245
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84476.5	1.1				ug/L	74699	Standard
	Be	9	178.3	18.7	**0.1869**	0.035	18.6	ug/L	8	Standard
	Al	27	1736036.7	3.1	**10.4319**	0.283	2.7	ug/L	1493	Standard
	Sc	45	72059.3	1.7				ug/L	61029	Standard
	Ti	47	2123.5	6.6	**4.6249**	0.308	6.7	ug/L	35	Standard
	V	51	32734.7	3.8	**3.4666**	0.135	3.9	ug/L	1435	Standard
	Cr	52	51911.1	1.7	**4.8949**	0.101	2.1	ug/L	7635	Standard
	Cr	53	5774.4	3.1	**4.9492**	0.156	3.2	ug/L	363	Standard
	Mn	55	492914.2	2.0	**59.6732**	1.219	2.0	ug/L	747	Standard
	Co	59	9193.8	6.0	**1.1391**	0.070	6.1	ug/L	234	Standard
	Ni	60	6665.5	3.0	**2.8840**	0.083	2.9	ug/L	215	Standard
	Cu	65	17774.1	0.7	**7.9654**	0.025	0.3	ug/L	104	Standard
	Zn	66	87339.7	0.8	**67.1049**	0.286	0.4	ug/L	363	Standard
>	**Ge**	72	585998.9	0.5				ug/L	495142	Standard
	As	75	970.2	5.1	**0.7914**	0.035	4.4	ug/L	-50	Standard
	Se	82	25.1	35.6	**0.0547**	0.068	124.2	ug/L	14	Standard
	Se-1	77	82.0	5.6	**0.3262**	0.050	15.4	ug/L	42	Standard
>	**Ga**	71	2373.5	5.4				mg/L	42	Standard
	Rb	85	23598.4	1.4				ug/L	17	Standard
	Y	89	515408.4	0.8				ug/L	417098	Standard
>	**Rh**	103	21.7	35.3				ug/L	13	Standard
	Mo	98	369.9	37.1	**0.1174**	0.046	39.0	ug/L	11	Standard
	Ag	107	699.0	38.5	**0.0822**	0.037	44.7	ug/L	68	Standard
	Cd	111	3080.1	3.0	**1.2985**	0.040	3.1	mg/L	2	Standard
	Cd	114	7389.1	5.7	**1.2756**	0.072	5.6	ug/L	12	Standard
>	**In**	115	662403.8	0.2				ug/L	587441	Standard
	Sn	118	660.0	13.5	**0.0055**	0.013	245.5	ug/L	497	Standard
	Sb	123	2141.2	12.8	**0.4025**	0.052	13.0	ug/L	78	Standard
	Ba	135	63716.0	0.4	**27.5914**	0.127	0.5	ug/L	24	Standard
	Ce	140	193219.3	1.3				ug/L	25	Standard
>	**Tb**	159	1027182.9	1.4				ug/L	939701	Standard
	Ho	165	2611.9	6.2				ug/L	10	Standard
	Tl	203	353.7	43.7	**0.0374**	0.017	46.4	ug/L	4	Standard
	Tl	205	421.7	67.1	**0.0541**	0.036	67.5	ug/L	2	Standard
	Pb	206	67035.7	2.8	**11.3788**	0.381	3.4	ug/L	240	Standard
	Pb	207	53097.6	2.6	**10.1377**	0.324	3.2	ug/L	198	Standard
	Pb	208	229986.8	1.7	**11.1165**	0.256	2.3	ug/L	883	Standard
	U	238	1351.4	18.8	**0.1748**	0.035	19.9	ug/L	3	Standard
>	**Bi**	209	574350.4	0.6				ug/L	523984	Standard

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:34:10
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	130.0	13.3	**0.0772**	0.020	26.0	mg/L	37	Standard
K	39	128.3	6.0	**0.1263**	0.018	14.2	mg/L	38	Standard
Ca	43	216.7	23.7	**9.0911**	2.889	31.8	mg/L	10	Standard
Fe	54	2661.6	0.9	**1.7733**	0.039	2.2	mg/L	179	Standard
Fe	57	965.0	5.3	**1.7027**	0.101	5.9	mg/L	335	Standard
> **Sc-1**	45	72059.3	1.7				mg/L	61029	Standard
Cl	35	30260.9	1.5				ug/L	20195	Standard
Kr	83	0.3	173.2				ug/L	4	Standard
Br	81	2210.2	10.2				ug/L	1547	Standard
P	31	20780.9	2.4				ug/L	15286	Standard
S	34	2430.2	2.3				ug/L	1572	Standard
Sr	88	83.3	13.9				ug/L	73	Standard
C	12	56.7	40.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	33.3	17.3				mg/L	3	Standard
Dy	164	4484.9	5.6				mg/L	6	Standard
Ho-1	165	2611.9	6.2				mg/L	10	Standard
Er	166	2556.9	13.6				mg/L	13	Standard
I	127	10483.6	6.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.090	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		118.350	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:34:10
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.761
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.612
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605SDL WG574690-02
Sample Date/Time: Wednesday, June 29, 2016 18:35:04
Number of Replicates: 3
Autosampler Position: 246
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83239.9	3.6				ug/L	74699	Standard
	Be	9	38.3	32.8	**0.0467**	0.013	27.9	ug/L	8	Standard
	Al	27	353337.7	1.0	**2.1460**	0.078	3.6	ug/L	1493	Standard
	Sc	45	70061.7	2.7				ug/L	61029	Standard
	Ti	47	444.3	5.4	**0.9019**	0.065	7.2	ug/L	35	Standard
	V	51	7638.3	1.2	**0.6761**	0.009	1.3	ug/L	1435	Standard
	Cr	52	17382.6	1.1	**0.9981**	0.034	3.5	ug/L	7635	Standard
	Cr	53	1533.4	2.6	**1.0999**	0.009	0.8	ug/L	363	Standard
	Mn	55	100723.4	1.5	**12.3208**	0.101	0.8	ug/L	747	Standard
	Co	59	1988.5	0.8	**0.2223**	0.008	3.5	ug/L	234	Standard
	Ni	60	1493.7	0.8	**0.5638**	0.012	2.1	ug/L	215	Standard
	Cu	65	3674.4	3.3	**1.5981**	0.060	3.8	ug/L	104	Standard
	Zn	66	19238.9	1.8	**14.7255**	0.173	1.2	ug/L	363	Standard
>	**Ge**	72	574541.2	2.2				ug/L	495142	Standard
	As	75	156.2	2.1	**0.1799**	0.005	2.7	ug/L	-50	Standard
	Se	82	23.9	29.5	**0.0491**	0.053	108.8	ug/L	14	Standard
	Se-1	77	58.0	21.2	**0.0603**	0.130	215.1	ug/L	42	Standard
>	**Ga**	71	518.3	4.3				mg/L	42	Standard
	Rb	85	4807.4	2.9				ug/L	17	Standard
	Y	89	501442.7	1.9				ug/L	417098	Standard
>	**Rh**	103	25.0	20.0				ug/L	13	Standard
	Mo	98	84.6	15.3	**0.0217**	0.004	18.8	ug/L	11	Standard
	Ag	107	254.3	9.6	**0.0208**	0.003	15.4	ug/L	68	Standard
	Cd	111	620.8	2.0	**0.2487**	0.004	1.5	mg/L	2	Standard
	Cd	114	1552.7	9.2	**0.2596**	0.025	9.8	ug/L	12	Standard
>	**In**	115	679737.8	0.5				ug/L	587441	Standard
	Sn	118	525.0	13.7	**-0.0171**	0.010	60.3	ug/L	497	Standard
	Sb	123	593.1	26.9	**0.1056**	0.030	28.3	ug/L	78	Standard
	Ba	135	12802.7	1.3	**5.3901**	0.099	1.8	ug/L	24	Standard
	Ce	140	39412.9	3.5				ug/L	25	Standard
>	**Tb**	159	1032176.9	1.0				ug/L	939701	Standard
	Ho	165	550.0	19.7				ug/L	10	Standard
	Tl	203	66.7	15.4	**0.0056**	0.001	20.8	ug/L	4	Standard
	Tl	205	78.3	41.0	**0.0102**	0.004	40.5	ug/L	2	Standard
	Pb	206	13522.4	0.6	**2.2309**	0.023	1.0	ug/L	240	Standard
	Pb	207	10801.2	2.0	**1.9994**	0.022	1.1	ug/L	198	Standard
	Pb	208	47468.1	0.8	**2.2283**	0.042	1.9	ug/L	883	Standard
	U	238	258.3	8.4	**0.0297**	0.002	8.2	ug/L	3	Standard
>	**Bi**	209	576211.7	1.3				ug/L	523984	Standard

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:37:21
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	16.4	**-0.0125**	0.009	74.1	mg/L	37	Standard
K	39	58.3	35.7	**-0.0022**	0.039	1787.9	mg/L	38	Standard
Ca	43	95.0	32.0	**2.7333**	1.777	65.0	mg/L	10	Standard
Fe	54	697.3	2.7	**0.3769**	0.020	5.3	mg/L	179	Standard
Fe	57	380.0	7.9	**0.0519**	0.104	201.0	mg/L	335	Standard
Sc-1	45	70061.7	2.7				mg/L	61029	Standard
Cl	35	30787.9	1.4				ug/L	20195	Standard
Kr	83	2.3	107.9				ug/L	4	Standard
Br	81	2130.1	1.6				ug/L	1547	Standard
P	31	20966.2	4.2				ug/L	15286	Standard
S	34	2485.2	3.8				ug/L	1572	Standard
Sr	88	90.0	33.3				ug/L	73	Standard
C	12	40.0	66.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	3	Standard
Dy	164	970.4	12.1				mg/L	6	Standard
Ho-1	165	550.0	19.7				mg/L	10	Standard
Er	166	480.0	14.6				mg/L	13	Standard
I	127	3620.4	13.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.434	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		116.036	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:37:21
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		115.712
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.967
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 18:38:18
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82768.9	2.2				ug/L	74699	Standard
	Be	9	55586.8	2.9	**56.9851**	0.437	0.8	ug/L	8	Standard
	Al	27	8619622.5	1.7	**52.9268**	0.666	1.3	ug/L	1493	Standard
	Sc	45	69535.9	1.6				ug/L	61029	Standard
	Ti	47	45959.0	2.5	**108.3957**	2.700	2.5	ug/L	35	Standard
	V	51	452339.2	1.9	**53.3985**	1.095	2.1	ug/L	1435	Standard
	Cr	52	452917.5	1.6	**53.6621**	0.813	1.5	ug/L	7635	Standard
	Cr	53	54972.9	2.4	**52.8966**	1.385	2.6	ug/L	363	Standard
	Mn	55	414134.0	1.2	**53.0785**	0.725	1.4	ug/L	747	Standard
	Co	59	411024.9	0.5	**55.6590**	0.257	0.5	ug/L	234	Standard
	Ni	60	111206.0	1.9	**53.0242**	0.984	1.9	ug/L	215	Standard
	Cu	65	110023.1	1.1	**52.7940**	0.716	1.4	ug/L	104	Standard
	Zn	66	64745.4	0.8	**52.5833**	0.259	0.5	ug/L	363	Standard
>	**Ge**	72	553341.4	0.3				ug/L	495142	Standard
	As	75	65551.1	0.4	**52.4265**	0.227	0.4	ug/L	-50	Standard
	Se	82	6664.4	1.0	**54.1963**	0.405	0.7	ug/L	14	Standard
	Se-1	77	4501.0	1.7	**54.5222**	0.959	1.8	ug/L	42	Standard
>	**Ga**	71	63.3	38.9				mg/L	42	Standard
	Rb	85	948.4	4.5				ug/L	17	Standard
	Y	89	490323.1	1.2				ug/L	417098	Standard
>	**Rh**	103	18.3	41.7				ug/L	13	Standard
	Mo	98	304786.5	0.3	**99.6729**	0.315	0.3	ug/L	11	Standard
	Ag	107	371315.6	1.3	**49.6520**	0.733	1.5	ug/L	68	Standard
	Cd	111	119443.2	1.0	**49.7891**	0.404	0.8	mg/L	2	Standard
	Cd	114	290803.0	0.5	**49.4253**	0.213	0.4	ug/L	12	Standard
>	**In**	115	673907.8	0.2				ug/L	587441	Standard
	Sn	118	339134.4	2.1	**50.5123**	1.079	2.1	ug/L	497	Standard
	Sb	123	262026.6	0.8	**48.9184**	0.455	0.9	ug/L	78	Standard
	Ba	135	111822.5	0.7	**47.6088**	0.434	0.9	ug/L	24	Standard
	Ce	140	206.7	14.6				ug/L	25	Standard
>	**Tb**	159	1024624.4	0.5				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	457896.5	0.3	**51.5737**	0.953	1.8	ug/L	4	Standard
	Tl	205	420019.2	1.0	**54.5041**	0.795	1.5	ug/L	2	Standard
	Pb	206	290158.6	1.2	**50.3552**	0.541	1.1	ug/L	240	Standard
	Pb	207	256755.7	1.2	**50.1530**	0.630	1.3	ug/L	198	Standard
	Pb	208	1071548.3	0.3	**52.9883**	1.261	2.4	ug/L	883	Standard
	U	238	389229.1	0.1	**52.5000**	1.135	2.2	ug/L	3	Standard
>	**Bi**	209	564495.9	2.1				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 18:40:35
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.7614**	29.023	173.2	mg/L	0	Standard
Mg	24	4642.4	0.2	**5.1955**	0.100	1.9	mg/L	37	Standard
K	39	2168.5	2.0	**4.0939**	0.144	3.5	mg/L	38	Standard
Ca	43	86.7	35.3	**2.2659**	1.624	71.7	mg/L	10	Standard
Fe	54	7113.4	2.2	**5.1545**	0.030	0.6	mg/L	179	Standard
Fe	57	2041.8	2.3	**5.0193**	0.241	4.8	mg/L	335	Standard
Sc-1	45	69535.9	1.6				mg/L	61029	Standard
Cl	35	30041.8	1.0				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	1843.4	4.1				ug/L	1547	Standard
P	31	21683.8	2.5				ug/L	15286	Standard
S	34	2823.6	5.0				ug/L	1572	Standard
Sr	88	53.3	39.0				ug/L	73	Standard
C	12	53.3	78.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	12.9	179.7				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	2093.5	24.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	113.970		
Al	27	105.854		
Sc	45			
Ti	47	108.396		
V	51	106.797		
Cr	52	107.324		
Cr	53			
Mn	55	106.157		
Co	59	111.318		
Ni	60	106.048		
Cu	65	105.588		
Zn	66	105.167		
Ge	72		111.754	
As	75	104.853		
Se	82	108.393		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 18:40:35
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	99.673		
Ag	107	99.304		
Cd	111	99.578		
Cd	114			
In	115		114.719	
Sn	118	101.025		
Sb	123	97.837		
Ba	135	95.218		
Ce	140			
Tb	159			
Ho	165			
Tl	203	103.147		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.977		
U	238	105.000		
Bi	209		107.731	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 18:40:35
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 18:41:29
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83048.8	2.8				ug/L	74699	Standard
	Be	9	23.3	68.9	**0.0313**	0.016	52.2	ug/L	8	Standard
	Al	27	3078.7	25.1	**0.0054**	0.004	78.6	ug/L	1493	Standard
	Sc	45	71367.7	1.6				ug/L	61029	Standard
	Ti	47	35.3	33.5	**-0.0277**	0.025	91.3	ug/L	35	Standard
	V	51	1419.9	5.2	**-0.0309**	0.006	18.5	ug/L	1435	Standard
	Cr	52	7844.0	0.2	**-0.0958**	0.021	21.7	ug/L	7635	Standard
	Cr	53	443.3	5.7	**0.0910**	0.034	37.3	ug/L	363	Standard
	Mn	55	990.7	12.0	**-0.0236**	0.012	50.6	ug/L	747	Standard
	Co	59	281.0	9.0	**-0.0000**	0.003	39793.0	ug/L	234	Standard
	Ni	60	196.7	7.3	**-0.0319**	0.005	14.4	ug/L	215	Standard
	Cu	65	145.3	12.4	**-0.0353**	0.007	21.2	ug/L	104	Standard
	Zn	66	464.0	9.4	**-0.0826**	0.026	32.0	ug/L	363	Standard
>	**Ge**	72	564975.6	2.4				ug/L	495142	Standard
	As	75	-39.3	63.0	**0.0286**	0.020	68.3	ug/L	-50	Standard
	Se	82	26.6	20.4	**0.0737**	0.038	51.8	ug/L	14	Standard
	Se-1	77	58.7	15.8	**0.0805**	0.100	124.7	ug/L	42	Standard
>	**Ga**	71	41.7	25.0				mg/L	42	Standard
	Rb	85	35.0	75.6				ug/L	17	Standard
	Y	89	486175.7	0.9				ug/L	417098	Standard
>	**Rh**	103	21.7	58.1				ug/L	13	Standard
	Mo	98	275.2	23.2	**0.0821**	0.020	24.6	ug/L	11	Standard
	Ag	107	144.3	16.7	**0.0059**	0.003	51.4	ug/L	68	Standard
	Cd	111	18.2	29.8	**-0.0005**	0.002	416.8	mg/L	2	Standard
	Cd	114	40.0	61.5	**0.0045**	0.004	89.7	ug/L	12	Standard
>	**In**	115	690622.1	0.6				ug/L	587441	Standard
	Sn	118	806.7	10.4	**0.0228**	0.013	56.6	ug/L	497	Standard
	Sb	123	553.4	29.9	**0.0967**	0.030	31.3	ug/L	78	Standard
	Ba	135	70.0	42.9	**0.0130**	0.012	94.6	ug/L	24	Standard
	Ce	140	61.7	54.0				ug/L	25	Standard
>	**Tb**	159	1047202.6	1.6				ug/L	939701	Standard
	Ho	165	16.7	34.6				ug/L	10	Standard
	Tl	203	75.3	48.5	**0.0063**	0.004	58.9	ug/L	4	Standard
	Tl	205	73.3	27.6	**0.0093**	0.002	26.3	ug/L	2	Standard
	Pb	206	333.0	14.3	**-0.0156**	0.007	42.7	ug/L	240	Standard
	Pb	207	293.0	14.2	**-0.0150**	0.006	43.1	ug/L	198	Standard
	Pb	208	1234.7	14.5	**-0.0154**	0.007	47.2	ug/L	883	Standard
	U	238	61.3	33.0	**0.0035**	0.002	69.2	ug/L	3	Standard
>	**Bi**	209	587080.9	2.4				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:43:46
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	38.3	15.1	**-0.0227**	0.007	30.4	mg/L	37	Standard
K	39	33.3	22.9	**-0.0512**	0.014	27.0	mg/L	38	Standard
Ca	43	28.3	50.9	**-0.9868**	0.802	81.3	mg/L	10	Standard
Fe	54	132.7	11.9	**-0.0421**	0.011	25.9	mg/L	179	Standard
Fe	57	280.0	10.7	**-0.2593**	0.100	38.4	mg/L	335	Standard
Sc-1	45	71367.7	1.6				mg/L	61029	Standard
Cl	35	29768.6	1.4				ug/L	20195	Standard
Kr	83	1.3	173.2				ug/L	4	Standard
Br	81	2106.8	7.7				ug/L	1547	Standard
P	31	21775.6	2.2				ug/L	15286	Standard
S	34	2523.5	3.0				ug/L	1572	Standard
Sr	88	70.0	49.5				ug/L	73	Standard
C	12	66.7	48.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	19.2	48.3				mg/L	6	Standard
Ho-1	165	16.7	34.6				mg/L	10	Standard
Er	166	16.7	91.7				mg/L	13	Standard
I	127	1905.1	33.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.104	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:43:46
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.564
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		112.042
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:43:46
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120701
Sample Date/Time: Wednesday, June 29, 2016 18:44:42
Number of Replicates: 3
Autosampler Position: 247
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82529.2	0.7				ug/L	74699	Standard
	Be	9	748.4	11.5	**0.7773**	0.092	11.9	ug/L	8	Standard
	Al	27	5186124.5	1.1	**31.9282**	0.368	1.2	ug/L	1493	Standard
	Sc	45	73214.8	0.3				ug/L	61029	Standard
	Ti	47	12465.8	1.5	**28.7779**	0.271	0.9	ug/L	35	Standard
	V	51	254152.3	1.7	**29.3595**	0.344	1.2	ug/L	1435	Standard
	Cr	52	145419.2	1.0	**16.2106**	0.127	0.8	ug/L	7635	Standard
	Cr	53	17483.4	2.8	**16.2818**	0.208	1.3	ug/L	363	Standard
	Mn	55	3993207.5	1.0	**503.6001**	3.344	0.7	ug/L	747	Standard
	Co	59	68748.0	1.0	**9.1067**	0.083	0.9	ug/L	234	Standard
	Ni	60	27938.1	3.3	**12.9787**	0.214	1.6	ug/L	215	Standard
	Cu	65	24664.7	1.5	**11.5352**	0.050	0.4	ug/L	104	Standard
	Zn	66	148101.8	1.7	**118.6204**	0.288	0.2	ug/L	363	Standard
>	**Ge**	72	563786.2	1.7				ug/L	495142	Standard
	As	75	6453.5	1.8	**5.1196**	0.031	0.6	ug/L	-50	Standard
	Se	82	62.5	2.1	**0.3624**	0.012	3.2	ug/L	14	Standard
	Se-1	77	114.0	16.7	**0.7504**	0.247	32.9	ug/L	42	Standard
>	**Ga**	71	11409.3	2.4				mg/L	42	Standard
	Rb	85	132848.2	0.6				ug/L	17	Standard
	Y	89	719521.7	0.4				ug/L	417098	Standard
>	**Rh**	103	43.3	43.7				ug/L	13	Standard
	Mo	98	809.7	3.4	**0.2640**	0.011	4.2	ug/L	11	Standard
	Ag	107	5298.9	1.5	**0.7088**	0.004	0.5	ug/L	68	Standard
	Cd	111	3980.4	2.6	**1.6822**	0.055	3.3	mg/L	2	Standard
	Cd	114	9713.7	1.4	**1.6793**	0.030	1.8	ug/L	12	Standard
>	**In**	115	661793.4	1.1				ug/L	587441	Standard
	Sn	118	911.7	2.8	**0.0439**	0.005	10.9	ug/L	497	Standard
	Sb	123	350.1	17.1	**0.0624**	0.012	19.0	ug/L	78	Standard
	Ba	135	341983.5	0.6	**148.3035**	0.748	0.5	ug/L	24	Standard
	Ce	140	2710163.1	0.4				ug/L	25	Standard
>	**Tb**	159	1034801.7	0.1				ug/L	939701	Standard
	Ho	165	19319.0	1.7				ug/L	10	Standard
	Tl	203	872.7	2.5	**0.0967**	0.002	2.5	ug/L	4	Standard
	Tl	205	843.4	5.5	**0.1098**	0.006	5.6	ug/L	2	Standard
	Pb	206	283131.4	1.0	**49.2257**	0.472	1.0	ug/L	240	Standard
	Pb	207	230315.5	1.3	**45.0642**	0.519	1.2	ug/L	198	Standard
	Pb	208	999318.0	0.9	**49.4947**	0.507	1.0	ug/L	883	Standard
	U	238	4899.8	1.7	**0.6576**	0.010	1.6	ug/L	3	Standard
>	**Bi**	209	563366.5	0.2				ug/L	523984	Standard

Sample ID: L1606120701
Report Date/Time: Wednesday, June 29, 2016 18:46:59
Page 1

Approved: July 04, 2016

Na	23	5.0	0.0	**48.0405**	0.159	0.3	mg/L	0	Standard
Mg	24	875.0	7.7	**0.8766**	0.075	8.6	mg/L	37	Standard
K	39	536.7	5.1	**0.8747**	0.054	6.2	mg/L	38	Standard
Ca	43	571.7	16.7	**27.5670**	4.925	17.9	mg/L	10	Standard
Fe	54	14788.8	1.9	**10.3138**	0.228	2.2	mg/L	179	Standard
Fe	57	4073.9	8.9	**10.4697**	1.059	10.1	mg/L	335	Standard
> **Sc-1**	45	73214.8	0.3				mg/L	61029	Standard
Cl	35	28264.3	0.5				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	3023.6	7.8				ug/L	1547	Standard
P	31	20599.0	2.2				ug/L	15286	Standard
S	34	2595.2	1.7				ug/L	1572	Standard
Sr	88	203.3	16.4				ug/L	73	Standard
C	12	130.0	40.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	150.0	43.7				mg/L	3	Standard
Dy	164	31532.4	1.2				mg/L	6	Standard
Ho-1	165	19319.0	1.7				mg/L	10	Standard
Er	166	17720.4	2.1				mg/L	13	Standard
I	127	25842.0	0.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.483	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.863	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120701
Report Date/Time: Wednesday, June 29, 2016 18:46:59
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	112.657
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	107.516
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120701
Report Date/Time: Wednesday, June 29, 2016 18:46:59
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120702
Sample Date/Time: Wednesday, June 29, 2016 18:47:54
Number of Replicates: 3
Autosampler Position: 248
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85381.8	3.5				ug/L	74699	Standard
	Be	9	623.3	3.3	**0.6272**	0.018	2.8	ug/L	8	Standard
	Al	27	4620203.8	1.9	**27.5041**	0.647	2.4	ug/L	1493	Standard
	Sc	45	75576.5	3.2				ug/L	61029	Standard
	Ti	47	6206.6	0.5	**14.3105**	0.192	1.3	ug/L	35	Standard
	V	51	134133.3	1.8	**15.4410**	0.153	1.0	ug/L	1435	Standard
	Cr	52	96500.3	0.7	**10.4416**	0.108	1.0	ug/L	7635	Standard
	Cr	53	11224.2	2.7	**10.3669**	0.448	4.3	ug/L	363	Standard
	Mn	55	3577332.4	1.7	**452.2740**	8.609	1.9	ug/L	747	Standard
	Co	59	47345.9	1.4	**6.2757**	0.086	1.4	ug/L	234	Standard
	Ni	60	24689.4	1.0	**11.4869**	0.101	0.9	ug/L	215	Standard
	Cu	65	29948.9	2.1	**14.0640**	0.244	1.7	ug/L	104	Standard
	Zn	66	164514.2	1.7	**132.1464**	1.033	0.8	ug/L	363	Standard
>	**Ge**	72	562390.3	1.6				ug/L	495142	Standard
	As	75	2776.0	3.7	**2.2412**	0.056	2.5	ug/L	-50	Standard
	Se	82	106.3	15.9	**0.7139**	0.124	17.3	ug/L	14	Standard
	Se-1	77	132.3	11.0	**0.9743**	0.190	19.5	ug/L	42	Standard
>	**Ga**	71	12940.5	1.1				mg/L	42	Standard
	Rb	85	250383.8	1.2				ug/L	17	Standard
	Y	89	695657.2	0.8				ug/L	417098	Standard
>	**Rh**	103	35.0	62.3				ug/L	13	Standard
	Mo	98	1645.8	3.3	**0.5188**	0.023	4.4	ug/L	11	Standard
	Ag	107	671.7	7.7	**0.0746**	0.007	9.4	ug/L	68	Standard
	Cd	111	2017.5	2.8	**0.8116**	0.028	3.5	mg/L	2	Standard
	Cd	114	4822.6	1.0	**0.7964**	0.005	0.7	ug/L	12	Standard
>	**In**	115	691832.3	1.1				ug/L	587441	Standard
	Sn	118	616.7	6.1	**-0.0050**	0.006	126.3	ug/L	497	Standard
	Sb	123	493.2	8.1	**0.0855**	0.008	9.1	ug/L	78	Standard
	Ba	135	281871.8	0.7	**116.9321**	1.732	1.5	ug/L	24	Standard
	Ce	140	1017603.0	1.2				ug/L	25	Standard
>	**Tb**	159	1062512.9	1.9				ug/L	939701	Standard
	Ho	165	14416.9	1.8				ug/L	10	Standard
	Tl	203	1059.0	3.0	**0.1166**	0.005	4.1	ug/L	4	Standard
	Tl	205	896.7	14.9	**0.1158**	0.019	16.8	ug/L	2	Standard
	Pb	206	100858.6	1.2	**17.3191**	0.532	3.1	ug/L	240	Standard
	Pb	207	81475.8	1.0	**15.7413**	0.458	2.9	ug/L	198	Standard
	Pb	208	356757.6	0.9	**17.4487**	0.446	2.6	ug/L	883	Standard
	U	238	6683.8	3.5	**0.8902**	0.048	5.4	ug/L	3	Standard
>	**Bi**	209	569106.4	1.9				ug/L	523984	Standard

Sample ID: L1606120702
Report Date/Time: Wednesday, June 29, 2016 18:50:11
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**31.5301**	27.313	86.6	mg/L	0	Standard
Mg	24	745.0	7.0	**0.7121**	0.061	8.5	mg/L	37	Standard
K	39	773.4	6.0	**1.2683**	0.120	9.5	mg/L	38	Standard
Ca	43	171.7	3.4	**6.2441**	0.546	8.7	mg/L	10	Standard
Fe	54	10724.1	1.5	**7.2104**	0.318	4.4	mg/L	179	Standard
Fe	57	2831.9	1.0	**6.7023**	0.187	2.8	mg/L	335	Standard
> **Sc-1**	45	75576.5	3.2				mg/L	61029	Standard
Cl	35	27596.4	1.5				ug/L	20195	Standard
Kr	83	4.0	66.1				ug/L	4	Standard
Br	81	8432.3	2.7				ug/L	1547	Standard
P	31	20674.1	3.5				ug/L	15286	Standard
S	34	2360.2	7.8				ug/L	1572	Standard
Sr	88	120.0	18.2				ug/L	73	Standard
C	12	350.0	15.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	106.7	43.3				mg/L	3	Standard
Dy	164	23142.4	1.5				mg/L	6	Standard
Ho-1	165	14416.9	1.8				mg/L	10	Standard
Er	166	13422.6	4.3				mg/L	13	Standard
I	127	174464.7	3.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		114.301	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.582	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120702
Report Date/Time: Wednesday, June 29, 2016 18:50:11
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	117.770
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.611
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120702
Report Date/Time: Wednesday, June 29, 2016 18:50:11
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120801
Sample Date/Time: Wednesday, June 29, 2016 18:51:05
Number of Replicates: 3
Autosampler Position: 249
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83030.4	3.3				ug/L	74699	Standard
	Be	9	1115.0	11.0	**1.1453**	0.091	8.0	ug/L	8	Standard
	Al	27	17314833.4	0.9	**106.0421**	2.666	2.5	ug/L	1493	Standard
	Sc	45	77183.1	5.1				ug/L	61029	Standard
	Ti	47	40548.2	1.5	**95.7436**	2.793	2.9	ug/L	35	Standard
	V	51	251782.0	1.4	**29.6946**	1.634	5.5	ug/L	1435	Standard
	Cr	52	243737.7	2.0	**28.4388**	0.710	2.5	ug/L	7635	Standard
	Cr	53	29872.8	2.3	**28.6239**	0.690	2.4	ug/L	363	Standard
	Mn	55	3685844.9	1.5	**474.5740**	27.736	5.8	ug/L	747	Standard
	Co	59	66271.6	1.6	**8.9620**	0.528	5.9	ug/L	234	Standard
	Ni	60	46026.2	2.6	**21.8948**	0.433	2.0	ug/L	215	Standard
	Cu	65	42002.8	1.5	**20.1164**	0.608	3.0	ug/L	104	Standard
	Zn	66	111824.0	2.6	**91.2379**	1.710	1.9	ug/L	363	Standard
>	**Ge**	72	553103.7	4.4				ug/L	495142	Standard
	As	75	5557.5	3.1	**4.5033**	0.093	2.1	ug/L	-50	Standard
	Se	82	102.2	5.1	**0.6977**	0.072	10.3	ug/L	14	Standard
	Se-1	77	134.3	12.1	**1.0314**	0.270	26.2	ug/L	42	Standard
>	**Ga**	71	24266.1	1.0				mg/L	42	Standard
	Rb	85	181545.0	2.7				ug/L	17	Standard
	Y	89	698995.8	3.0				ug/L	417098	Standard
>	**Rh**	103	88.3	6.5				ug/L	13	Standard
	Mo	98	2397.1	2.7	**0.8492**	0.049	5.8	ug/L	11	Standard
	Ag	107	1012.7	3.0	**0.1348**	0.009	6.5	ug/L	68	Standard
	Cd	111	911.2	3.1	**0.4060**	0.012	2.9	mg/L	2	Standard
	Cd	114	2080.2	4.3	**0.3830**	0.011	2.9	ug/L	12	Standard
>	**In**	115	618664.5	3.1				ug/L	587441	Standard
	Sn	118	1240.1	3.1	**0.1071**	0.013	12.0	ug/L	497	Standard
	Sb	123	1640.7	2.7	**0.3299**	0.019	5.8	ug/L	78	Standard
	Ba	135	534847.9	1.0	**248.3132**	9.882	4.0	ug/L	24	Standard
	Ce	140	1227932.9	1.4				ug/L	25	Standard
>	**Tb**	159	991058.9	3.3				ug/L	939701	Standard
	Ho	165	19152.1	0.5				ug/L	10	Standard
	Tl	203	1084.7	2.8	**0.1294**	0.004	2.8	ug/L	4	Standard
	Tl	205	976.7	6.4	**0.1361**	0.004	2.8	ug/L	2	Standard
	Pb	206	122644.8	1.6	**22.8364**	1.340	5.9	ug/L	240	Standard
	Pb	207	98294.5	1.0	**20.5914**	1.111	5.4	ug/L	198	Standard
	Pb	208	423360.3	0.4	**22.4444**	1.000	4.5	ug/L	883	Standard
	U	238	7556.2	1.8	**1.0902**	0.045	4.1	ug/L	3	Standard
>	**Bi**	209	526041.8	4.3				ug/L	523984	Standard

Sample ID: L1606120801
Report Date/Time: Wednesday, June 29, 2016 18:53:22
Page 1

Approved: July 04, 2016

Na	23	10.0	50.0	**92.8331**	50.540	54.4	mg/L	0	Standard
Mg	24	373.3	13.0	**0.3164**	0.050	15.9	mg/L	37	Standard
K	39	846.7	3.0	**1.3673**	0.064	4.7	mg/L	38	Standard
Ca	43	1881.8	8.7	**91.3781**	4.657	5.1	mg/L	10	Standard
Fe	54	16605.4	3.8	**11.0236**	0.942	8.5	mg/L	179	Standard
Fe	57	5247.6	6.9	**13.0290**	0.645	4.9	mg/L	335	Standard
> **Sc-1**	45	77183.1	5.1				mg/L	61029	Standard
Cl	35	28165.5	0.9				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	3540.4	14.7				ug/L	1547	Standard
P	31	23239.5	5.6				ug/L	15286	Standard
S	34	2426.9	3.5				ug/L	1572	Standard
Sr	88	415.0	11.5				ug/L	73	Standard
C	12	173.3	6.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	66.7	60.6				mg/L	3	Standard
Dy	164	30988.1	3.0				mg/L	6	Standard
Ho-1	165	19152.1	0.5				mg/L	10	Standard
Er	166	17717.0	2.5				mg/L	13	Standard
I	127	27493.3	1.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.154	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.706	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120801
Report Date/Time: Wednesday, June 29, 2016 18:53:22
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	105.315	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	100.393	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120802
Sample Date/Time: Wednesday, June 29, 2016 18:54:17
Number of Replicates: 3
Autosampler Position: 250
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88923.6	2.8				ug/L	74699	Standard
	Be	9	661.7	11.4	**0.6387**	0.067	10.5	ug/L	8	Standard
	Al	27	15492795.6	1.0	**88.5762**	1.812	2.0	ug/L	1493	Standard
	Sc	45	80386.0	1.0				ug/L	61029	Standard
	Ti	47	20524.2	1.6	**45.7756**	1.470	3.2	ug/L	35	Standard
	V	51	220306.5	1.3	**24.5176**	0.534	2.2	ug/L	1435	Standard
	Cr	52	643370.5	1.6	**72.5457**	2.733	3.8	ug/L	7635	Standard
	Cr	53	78035.6	3.0	**71.2237**	3.200	4.5	ug/L	363	Standard
	Mn	55	3332205.3	1.0	**405.3306**	7.931	2.0	ug/L	747	Standard
	Co	59	49008.5	0.3	**6.2510**	0.159	2.5	ug/L	234	Standard
	Ni	60	36936.6	2.2	**16.5952**	0.689	4.2	ug/L	215	Standard
	Cu	65	48506.2	1.8	**21.9786**	0.680	3.1	ug/L	104	Standard
	Zn	66	431997.7	0.4	**334.6356**	8.273	2.5	ug/L	363	Standard
>	**Ge**	72	584593.6	2.3				ug/L	495142	Standard
	As	75	4706.8	0.7	**3.6196**	0.063	1.8	ug/L	-50	Standard
	Se	82	170.0	3.4	**1.1737**	0.017	1.5	ug/L	14	Standard
	Se-1	77	180.7	11.7	**1.4743**	0.259	17.6	ug/L	42	Standard
>	**Ga**	71	19881.4	1.3				mg/L	42	Standard
	Rb	85	200103.0	1.8				ug/L	17	Standard
	Y	89	657665.4	0.6				ug/L	417098	Standard
>	**Rh**	103	171.7	7.3				ug/L	13	Standard
	Mo	98	1216.9	1.5	**0.3934**	0.011	2.8	ug/L	11	Standard
	Ag	107	1006.0	3.7	**0.1221**	0.007	5.4	ug/L	68	Standard
	Cd	111	4519.9	2.1	**1.8818**	0.049	2.6	mg/L	2	Standard
	Cd	114	11256.2	1.6	**1.9163**	0.030	1.6	ug/L	12	Standard
>	**In**	115	672127.4	1.5				ug/L	587441	Standard
	Sn	118	1105.0	15.3	**0.0707**	0.026	36.7	ug/L	497	Standard
	Sb	123	229.1	10.3	**0.0387**	0.005	12.5	ug/L	78	Standard
	Ba	135	361495.2	1.1	**154.3886**	3.928	2.5	ug/L	24	Standard
	Ce	140	936346.8	0.9				ug/L	25	Standard
>	**Tb**	159	1070983.4	3.2				ug/L	939701	Standard
	Ho	165	13711.2	2.1				ug/L	10	Standard
	Tl	203	1023.7	3.8	**0.1103**	0.006	5.8	ug/L	4	Standard
	Tl	205	995.0	4.8	**0.1256**	0.007	5.2	ug/L	2	Standard
	Pb	206	1451242.4	0.5	**244.9779**	5.608	2.3	ug/L	240	Standard
	Pb	207	1189055.7	0.7	**225.9168**	5.756	2.5	ug/L	198	Standard
	Pb	208	5216722.9	1.2	**250.8847**	7.037	2.8	ug/L	883	Standard
	U	238	9013.7	1.2	**1.1768**	0.034	2.9	ug/L	3	Standard
>	**Bi**	209	581076.6	1.9				ug/L	523984	Standard

Sample ID: L1606120802
Report Date/Time: Wednesday, June 29, 2016 18:56:34
Page 1

Approved: July 04, 2016

Na	23	6.7	173.2	**57.8631**	100.213	173.2	mg/L	0	Standard
Mg	24	553.3	14.7	**0.4776**	0.083	17.4	mg/L	37	Standard
K	39	1005.0	9.9	**1.5724**	0.167	10.6	mg/L	38	Standard
Ca	43	1240.1	6.3	**56.9237**	3.160	5.6	mg/L	10	Standard
Fe	54	17498.7	0.5	**11.1262**	0.159	1.4	mg/L	179	Standard
Fe	57	5124.2	2.5	**12.1481**	0.285	2.3	mg/L	335	Standard
Sc-1	45	80386.0	1.0				mg/L	61029	Standard
Cl	35	28790.7	0.3				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	8875.9	6.3				ug/L	1547	Standard
P	31	20292.0	8.4				ug/L	15286	Standard
S	34	2478.5	4.1				ug/L	1572	Standard
Sr	88	438.3	2.4				ug/L	73	Standard
C	12	323.3	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	136.7	34.6				mg/L	3	Standard
Dy	164	22618.6	1.5				mg/L	6	Standard
Ho-1	165	13711.2	2.1				mg/L	10	Standard
Er	166	12578.6	1.5				mg/L	13	Standard
I	127	73652.8	8.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		119.043	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		118.066	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120802
Report Date/Time: Wednesday, June 29, 2016 18:56:34
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.416
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.896
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120802
Report Date/Time: Wednesday, June 29, 2016 18:56:34
Page 3

Approved: July 04, 2016

Pb 206 Upper, S, EEE	Pb	206
Pb 207 Upper, S, EEE	Pb	207
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129101
Sample Date/Time: Wednesday, June 29, 2016 18:57:28
Number of Replicates: 3
Autosampler Position: 251
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89580.8	1.7				ug/L	74699	Standard
	Be	9	360.0	14.7	**0.3489**	0.053	15.3	ug/L	8	Standard
	Al	27	3187676.8	1.0	**18.0772**	0.341	1.9	ug/L	1493	Standard
	Sc	45	73445.9	1.6				ug/L	61029	Standard
	Ti	47	23807.7	6.7	**54.3513**	2.712	5.0	ug/L	35	Standard
	V	51	62405.7	2.9	**6.9715**	0.072	1.0	ug/L	1435	Standard
	Cr	52	100378.5	1.9	**10.7261**	0.052	0.5	ug/L	7635	Standard
	Cr	53	11571.1	3.7	**10.5314**	0.357	3.4	ug/L	363	Standard
	Mn	55	3995310.3	0.7	**497.7426**	6.886	1.4	ug/L	747	Standard
	Co	59	12689.3	1.5	**1.6298**	0.018	1.1	ug/L	234	Standard
	Ni	60	15702.5	1.8	**7.1516**	0.026	0.4	ug/L	215	Standard
	Cu	65	1030601.6	2.5	**480.2115**	2.610	0.5	ug/L	104	Standard
	Zn	66	647092.9	2.6	**513.3926**	3.363	0.7	ug/L	363	Standard
>	**Ge**	72	570780.0	2.0				ug/L	495142	Standard
	As	75	23305.2	1.3	**18.1104**	0.138	0.8	ug/L	-50	Standard
	Se	82	481.4	2.5	**3.6678**	0.121	3.3	ug/L	14	Standard
	Se-1	77	377.3	9.1	**3.8600**	0.402	10.4	ug/L	42	Standard
>	**Ga**	71	43582.7	4.5				mg/L	42	Standard
	Rb	85	26201.0	3.0				ug/L	17	Standard
	Y	89	551000.5	1.0				ug/L	417098	Standard
>	**Rh**	103	178.3	18.7				ug/L	13	Standard
	Mo	98	21134.4	1.3	**6.6476**	0.191	2.9	ug/L	11	Standard
	Ag	107	15204.6	2.3	**1.9441**	0.032	1.6	ug/L	68	Standard
	Cd	111	1466.1	5.3	**0.5801**	0.025	4.4	mg/L	2	Standard
	Cd	114	3712.7	3.5	**0.6050**	0.007	1.1	ug/L	12	Standard
>	**In**	115	700391.9	2.5				ug/L	587441	Standard
	Sn	118	18187.6	2.5	**2.5189**	0.131	5.2	ug/L	497	Standard
	Sb	123	7709.0	2.8	**1.3807**	0.022	1.6	ug/L	78	Standard
	Ba	135	971846.2	1.7	**398.2918**	4.112	1.0	ug/L	24	Standard
	Ce	140	138638.2	1.7				ug/L	25	Standard
>	**Tb**	159	1070571.9	3.6				ug/L	939701	Standard
	Ho	165	2720.2	8.8				ug/L	10	Standard
	Tl	203	657.3	30.6	**0.0422**	0.014	32.0	ug/L	4	Standard
	Tl	205	563.3	26.0	**0.0436**	0.011	26.1	ug/L	2	Standard
	Pb	206	89328.6	2.2	**9.1434**	0.147	1.6	ug/L	240	Standard
	Pb	207	72922.4	1.4	**8.3969**	0.107	1.3	ug/L	198	Standard
	Pb	208	316979.9	0.6	**9.2432**	0.217	2.4	ug/L	883	Standard
	U	238	78871.8	1.7	**6.3120**	0.222	3.5	ug/L	3	Standard
>	**Bi**	209	950965.0	2.0				ug/L	523984	Standard

Sample ID: L1606129101
Report Date/Time: Wednesday, June 29, 2016 18:59:45
Page 1

Approved: July 04, 2016

Na	23	5.0	173.2	**48.7719**	84.467	173.2	mg/L	0	Standard
Mg	24	613.3	9.8	**0.5934**	0.073	12.3	mg/L	37	Standard
K	39	323.3	7.3	**0.4798**	0.044	9.2	mg/L	38	Standard
Ca	43	198.3	14.8	**7.8949**	1.614	20.4	mg/L	10	Standard
Fe	54	8830.8	5.5	**6.0856**	0.392	6.4	mg/L	179	Standard
Fe	57	2530.2	2.9	**6.0730**	0.228	3.8	mg/L	335	Standard
Sc-1	45	73445.9	1.6				mg/L	61029	Standard
Cl	35	25952.9	0.3				ug/L	20195	Standard
Kr	83	2.0	100.0				ug/L	4	Standard
Br	81	8005.5	5.5				ug/L	1547	Standard
P	31	19060.3	1.8				ug/L	15286	Standard
S	34	1756.8	1.0				ug/L	1572	Standard
Sr	88	518.3	2.8				ug/L	73	Standard
C	12	1226.7	9.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	1270.1	11.0				mg/L	3	Standard
Dy	164	4157.3	8.4				mg/L	6	Standard
Ho-1	165	2720.2	8.8				mg/L	10	Standard
Er	166	2643.6	8.4				mg/L	13	Standard
I	127	191520.8	4.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		119.923	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.276	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606129101
Report Date/Time: Wednesday, June 29, 2016 18:59:45
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	119.228
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	181.487
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Cu 65 Upper, S, EEE	Cu	65	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135	
Bi 209 Int Std for sample	Bi	209	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138101
Sample Date/Time: Wednesday, June 29, 2016 19:00:40
Number of Replicates: 3
Autosampler Position: 252
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96098.8	1.2				ug/L	74699	Standard
	Be	9	1650.1	8.9	**1.4642**	0.124	8.5	ug/L	8	Standard
	Al	27	23502208.7	0.7	**124.3024**	1.056	0.8	ug/L	1493	Standard
	Sc	45	81510.4	1.4				ug/L	61029	Standard
	Ti	47	9357.2	0.5	**21.1737**	0.376	1.8	ug/L	35	Standard
	V	51	257600.4	1.3	**29.2098**	0.781	2.7	ug/L	1435	Standard
	Cr	52	224787.2	0.6	**25.1231**	0.400	1.6	ug/L	7635	Standard
	Cr	53	27600.1	3.0	**25.4059**	0.410	1.6	ug/L	363	Standard
	Mn	55	5543699.6	0.4	**686.2289**	8.924	1.3	ug/L	747	Standard
	Co	59	117640.6	0.9	**15.3216**	0.399	2.6	ug/L	234	Standard
	Ni	60	64249.3	0.8	**29.4562**	0.340	1.2	ug/L	215	Standard
	Cu	65	57365.7	1.4	**26.4629**	0.232	0.9	ug/L	104	Standard
	Zn	66	103406.0	1.2	**81.1457**	1.055	1.3	ug/L	363	Standard
>	**Ge**	72	574477.7	1.7				ug/L	495142	Standard
	As	75	12246.4	0.5	**9.4841**	0.110	1.2	ug/L	-50	Standard
	Se	82	125.1	3.5	**0.8446**	0.024	2.8	ug/L	14	Standard
	Se-1	77	164.3	2.5	**1.3176**	0.070	5.3	ug/L	42	Standard
>	**Ga**	71	29587.2	2.2				mg/L	42	Standard
	Rb	85	313937.1	2.6				ug/L	17	Standard
	Y	89	822636.0	0.8				ug/L	417098	Standard
>	**Rh**	103	38.3	49.4				ug/L	13	Standard
	Mo	98	757.2	4.8	**0.2454**	0.010	4.2	ug/L	11	Standard
	Ag	107	953.4	1.4	**0.1164**	0.003	2.3	ug/L	68	Standard
	Cd	111	1588.5	0.6	**0.6638**	0.009	1.4	mg/L	2	Standard
	Cd	114	3783.2	0.8	**0.6501**	0.009	1.4	ug/L	12	Standard
>	**In**	115	664448.0	1.0				ug/L	587441	Standard
	Sn	118	643.3	5.2	**0.0026**	0.004	152.2	ug/L	497	Standard
	Sb	123	269.8	13.5	**0.0469**	0.007	15.6	ug/L	78	Standard
	Ba	135	662128.9	0.5	**286.0118**	2.987	1.0	ug/L	24	Standard
	Ce	140	1687892.1	1.1				ug/L	25	Standard
>	**Tb**	159	1062435.8	0.8				ug/L	939701	Standard
	Ho	165	27182.7	2.2				ug/L	10	Standard
	Tl	203	1364.1	4.3	**0.1521**	0.007	4.3	ug/L	4	Standard
	Tl	205	1335.1	6.0	**0.1737**	0.010	5.7	ug/L	2	Standard
	Pb	206	284182.0	0.3	**49.3998**	0.214	0.4	ug/L	240	Standard
	Pb	207	234533.7	1.0	**45.8828**	0.448	1.0	ug/L	198	Standard
	Pb	208	1023299.4	0.1	**50.6751**	0.295	0.6	ug/L	883	Standard
	U	238	6312.3	2.1	**0.8483**	0.014	1.7	ug/L	3	Standard
>	**Bi**	209	563477.4	0.7				ug/L	523984	Standard

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:02:57
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**28.8214**	24.963	86.6	mg/L	0	Standard
Mg	24	151.7	18.2	**0.0816**	0.027	33.5	mg/L	37	Standard
K	39	1206.7	9.1	**1.8839**	0.198	10.5	mg/L	38	Standard
Ca	43	230.0	38.1	**8.3188**	4.002	48.1	mg/L	10	Standard
Fe	54	27991.0	1.6	**17.6350**	0.523	3.0	mg/L	179	Standard
Fe	57	7366.8	1.7	**17.6771**	0.511	2.9	mg/L	335	Standard
Sc-1	45	81510.4	1.4				mg/L	61029	Standard
Cl	35	28359.2	2.3				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	4510.7	3.7				ug/L	1547	Standard
P	31	20567.3	1.1				ug/L	15286	Standard
S	34	2398.5	2.1				ug/L	1572	Standard
Sr	88	138.3	12.7				ug/L	73	Standard
C	12	220.0	4.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	113.3	35.7				mg/L	3	Standard
Dy	164	44981.5	2.0				mg/L	6	Standard
Ho-1	165	27182.7	2.2				mg/L	10	Standard
Er	166	24606.7	4.9				mg/L	13	Standard
I	127	24668.4	3.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		128.648	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		116.023	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:02:57
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.109
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.537
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:02:57
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138101
Sample Date/Time: Wednesday, June 29, 2016 19:03:51
Number of Replicates: 3
Autosampler Position: 253
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89738.3	1.4				ug/L	74699	Standard
	Be	9	41.7	6.9	**0.0470**	0.003	6.5	ug/L	8	Standard
	Al	27	505536.2	3.8	**2.8498**	0.087	3.1	ug/L	1493	Standard
	Sc	45	72476.3	1.0				ug/L	61029	Standard
	Ti	47	238.7	3.3	**0.4178**	0.015	3.5	ug/L	35	Standard
	V	51	6994.4	2.1	**0.5793**	0.021	3.7	ug/L	1435	Standard
	Cr	52	13067.6	0.8	**0.4515**	0.011	2.5	ug/L	7635	Standard
	Cr	53	1060.0	5.3	**0.6298**	0.052	8.3	ug/L	363	Standard
	Mn	55	117506.1	2.5	**13.9721**	0.301	2.2	ug/L	747	Standard
	Co	59	2633.2	1.9	**0.2964**	0.006	2.1	ug/L	234	Standard
	Ni	60	1579.1	3.6	**0.5819**	0.030	5.1	ug/L	215	Standard
	Cu	65	1415.4	1.9	**0.5327**	0.017	3.2	ug/L	104	Standard
	Zn	66	4415.0	0.5	**2.9272**	0.025	0.9	ug/L	363	Standard
>	**Ge**	72	591804.3	0.7				ug/L	495142	Standard
	As	75	188.6	12.3	**0.2004**	0.017	8.4	ug/L	-50	Standard
	Se	82	17.7	38.3	**-0.0030**	0.051	1669.1	ug/L	14	Standard
	Se-1	77	53.3	24.5	**-0.0120**	0.145	1206.4	ug/L	42	Standard
>	**Ga**	71	651.7	3.8				mg/L	42	Standard
	Rb	85	6646.5	2.9				ug/L	17	Standard
	Y	89	496582.4	1.4				ug/L	417098	Standard
>	**Rh**	103	10.0	86.6				ug/L	13	Standard
	Mo	98	63.7	13.0	**0.0149**	0.003	17.4	ug/L	11	Standard
	Ag	107	138.0	7.0	**0.0054**	0.001	21.7	ug/L	68	Standard
	Cd	111	41.6	25.2	**0.0093**	0.004	47.8	mg/L	2	Standard
	Cd	114	78.4	6.5	**0.0111**	0.001	8.0	ug/L	12	Standard
>	**In**	115	680653.4	0.9				ug/L	587441	Standard
	Sn	118	558.3	7.0	**-0.0122**	0.005	43.2	ug/L	497	Standard
	Sb	123	143.1	16.8	**0.0222**	0.004	19.8	ug/L	78	Standard
	Ba	135	14129.9	0.6	**5.9423**	0.035	0.6	ug/L	24	Standard
	Ce	140	36747.8	1.4				ug/L	25	Standard
>	**Tb**	159	1046560.0	0.5				ug/L	939701	Standard
	Ho	165	620.0	6.6				ug/L	10	Standard
	Tl	203	40.7	9.9	**0.0027**	0.000	16.9	ug/L	4	Standard
	Tl	205	38.3	19.9	**0.0050**	0.001	18.8	ug/L	2	Standard
	Pb	206	6390.7	0.3	**1.0069**	0.002	0.2	ug/L	240	Standard
	Pb	207	5305.6	1.7	**0.9374**	0.017	1.8	ug/L	198	Standard
	Pb	208	22358.8	1.4	**1.0005**	0.020	2.0	ug/L	883	Standard
	U	238	147.3	10.0	**0.0149**	0.002	12.5	ug/L	3	Standard
>	**Bi**	209	581458.5	0.4				ug/L	523984	Standard

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:06:08
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	36.7	41.7	**-0.0251**	0.017	67.5	mg/L	37	Standard
K	39	73.3	27.6	**0.0225**	0.038	170.4	mg/L	38	Standard
Ca	43	25.0	40.0	**-1.1927**	0.538	45.1	mg/L	10	Standard
Fe	54	777.1	4.9	**0.4166**	0.030	7.2	mg/L	179	Standard
Fe	57	423.3	1.8	**0.1373**	0.012	9.1	mg/L	335	Standard
> **Sc-1**	45	72476.3	1.0				mg/L	61029	Standard
Cl	35	26907.2	2.4				ug/L	20195	Standard
Kr	83	1.7	124.9				ug/L	4	Standard
Br	81	2140.2	2.1				ug/L	1547	Standard
P	31	19854.7	4.5				ug/L	15286	Standard
S	34	2031.8	8.8				ug/L	1572	Standard
Sr	88	61.7	24.8				ug/L	73	Standard
C	12	66.7	82.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	1031.0	18.9				mg/L	6	Standard
Ho-1	165	620.0	6.6				mg/L	10	Standard
Er	166	536.7	12.4				mg/L	13	Standard
I	127	3175.3	27.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		120.134	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		119.522	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:06:08
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		115.868
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.969
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138102
Sample Date/Time: Wednesday, June 29, 2016 19:07:03
Number of Replicates: 3
Autosampler Position: 254
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93009.7	1.6				ug/L	74699	Standard
	Be	9	1248.4	3.0	**1.1470**	0.051	4.5	ug/L	8	Standard
	Al	27	40124336.3	0.4	**219.2936**	2.793	1.3	ug/L	1493	Standard
	Sc	45	83437.5	1.7				ug/L	61029	Standard
	Ti	47	9733.8	0.5	**22.4436**	0.290	1.3	ug/L	35	Standard
	V	51	219254.4	2.3	**25.2954**	0.490	1.9	ug/L	1435	Standard
	Cr	52	167385.2	0.7	**18.8093**	0.105	0.6	ug/L	7635	Standard
	Cr	53	20240.2	3.5	**18.8976**	0.529	2.8	ug/L	363	Standard
	Mn	55	6543259.4	1.4	**825.1579**	15.855	1.9	ug/L	747	Standard
	Co	59	86492.3	2.3	**11.4640**	0.252	2.2	ug/L	234	Standard
	Ni	60	58233.4	1.3	**27.1864**	0.121	0.4	ug/L	215	Standard
	Cu	65	35144.1	1.8	**16.4764**	0.210	1.3	ug/L	104	Standard
	Zn	66	61921.6	1.7	**49.3206**	0.509	1.0	ug/L	363	Standard
>	**Ge**	72	563884.1	0.9				ug/L	495142	Standard
	As	75	5876.5	0.9	**4.6669**	0.080	1.7	ug/L	-50	Standard
	Se	82	62.7	20.2	**0.3638**	0.101	27.7	ug/L	14	Standard
	Se-1	77	116.3	8.0	**0.7757**	0.099	12.7	ug/L	42	Standard
>	**Ga**	71	24224.4	1.6				mg/L	42	Standard
	Rb	85	231886.4	1.9				ug/L	17	Standard
	Y	89	788603.8	0.2				ug/L	417098	Standard
>	**Rh**	103	76.7	50.7				ug/L	13	Standard
	Mo	98	590.2	6.7	**0.1909**	0.011	6.0	ug/L	11	Standard
	Ag	107	590.3	14.1	**0.0675**	0.010	15.2	ug/L	68	Standard
	Cd	111	402.0	10.7	**0.1627**	0.016	9.9	mg/L	2	Standard
	Cd	114	896.6	5.0	**0.1531**	0.007	4.5	ug/L	12	Standard
>	**In**	115	661468.7	1.5				ug/L	587441	Standard
	Sn	118	563.3	17.0	**-0.0090**	0.016	173.2	ug/L	497	Standard
	Sb	123	168.4	11.6	**0.0279**	0.004	15.0	ug/L	78	Standard
	Ba	135	766980.0	0.9	**332.7995**	3.011	0.9	ug/L	24	Standard
	Ce	140	1513996.2	0.9				ug/L	25	Standard
>	**Tb**	159	1057537.7	1.8				ug/L	939701	Standard
	Ho	165	25513.1	1.6				ug/L	10	Standard
	Tl	203	1321.7	3.8	**0.1489**	0.005	3.2	ug/L	4	Standard
	Tl	205	1201.7	6.5	**0.1579**	0.009	5.7	ug/L	2	Standard
	Pb	206	69846.7	2.1	**12.2126**	0.140	1.1	ug/L	240	Standard
	Pb	207	56032.1	2.8	**11.0205**	0.188	1.7	ug/L	198	Standard
	Pb	208	243026.5	0.6	**12.1032**	0.059	0.5	ug/L	883	Standard
	U	238	7981.8	0.7	**1.0851**	0.004	0.3	ug/L	3	Standard
>	**Bi**	209	557709.9	1.1				ug/L	523984	Standard

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:09:20
Page 1

Approved: July 04, 2016

Na	23	6.7	86.6	**56.4197**	48.874	86.6	mg/L	0	Standard
Mg	24	238.3	32.7	**0.1607**	0.077	48.1	mg/L	37	Standard
K	39	688.3	8.8	**0.9980**	0.083	8.3	mg/L	38	Standard
Ca	43	911.7	12.3	**39.6249**	5.390	13.6	mg/L	10	Standard
Fe	54	20800.6	0.9	**12.7625**	0.173	1.4	mg/L	179	Standard
Fe	57	6012.9	2.1	**13.8749**	0.248	1.8	mg/L	335	Standard
> **Sc-1**	45	83437.5	1.7				mg/L	61029	Standard
Cl	35	29332.4	1.9				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2603.6	11.5				ug/L	1547	Standard
P	31	21632.1	2.1				ug/L	15286	Standard
S	34	2746.9	3.7				ug/L	1572	Standard
Sr	88	253.3	10.9				ug/L	73	Standard
C	12	116.7	27.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	33.3	45.8				mg/L	3	Standard
Dy	164	42471.8	1.0				mg/L	6	Standard
Ho-1	165	25513.1	1.6				mg/L	10	Standard
Er	166	22212.9	0.4				mg/L	13	Standard
I	127	23259.5	0.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		124.513	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		113.883	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:09:20
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.602
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.436
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:09:20
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138102
Sample Date/Time: Wednesday, June 29, 2016 19:10:14
Number of Replicates: 3
Autosampler Position: 255
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85837.5	1.0				ug/L	74699	Standard
	Be	9	38.3	19.9	**0.0455**	0.007	16.1	ug/L	8	Standard
	Al	27	829500.0	3.1	**4.8979**	0.105	2.1	ug/L	1493	Standard
	Sc	45	72806.2	1.9				ug/L	61029	Standard
	Ti	47	228.0	6.2	**0.3999**	0.033	8.1	ug/L	35	Standard
	V	51	5937.2	2.5	**0.4694**	0.017	3.6	ug/L	1435	Standard
	Cr	52	11356.6	1.4	**0.2725**	0.020	7.2	ug/L	7635	Standard
	Cr	53	790.0	13.9	**0.3928**	0.096	24.5	ug/L	363	Standard
	Mn	55	140399.5	1.1	**16.9108**	0.252	1.5	ug/L	747	Standard
	Co	59	1986.1	4.2	**0.2171**	0.011	4.8	ug/L	234	Standard
	Ni	60	1398.7	1.4	**0.5081**	0.009	1.8	ug/L	215	Standard
	Cu	65	887.4	6.1	**0.2997**	0.027	8.9	ug/L	104	Standard
	Zn	66	3530.4	0.8	**2.2796**	0.035	1.5	ug/L	363	Standard
>	**Ge**	72	585343.3	0.6				ug/L	495142	Standard
	As	75	60.2	40.7	**0.1052**	0.019	17.9	ug/L	-50	Standard
	Se	82	19.5	23.7	**0.0127**	0.037	290.2	ug/L	14	Standard
	Se-1	77	54.0	19.3	**0.0032**	0.122	3775.1	ug/L	42	Standard
>	**Ga**	71	611.7	10.2				mg/L	42	Standard
	Rb	85	4939.1	4.6				ug/L	17	Standard
	Y	89	509652.0	1.0				ug/L	417098	Standard
>	**Rh**	103	5.0	100.0				ug/L	13	Standard
	Mo	98	44.1	7.7	**0.0084**	0.001	14.8	ug/L	11	Standard
	Ag	107	136.0	8.3	**0.0049**	0.002	33.6	ug/L	68	Standard
	Cd	111	13.6	15.2	**-0.0024**	0.001	32.7	mg/L	2	Standard
	Cd	114	34.6	39.9	**0.0036**	0.002	62.3	ug/L	12	Standard
>	**In**	115	689569.5	1.2				ug/L	587441	Standard
	Sn	118	635.0	8.5	**-0.0021**	0.008	369.1	ug/L	497	Standard
	Sb	123	120.9	7.4	**0.0179**	0.002	10.7	ug/L	78	Standard
	Ba	135	16312.1	1.7	**6.7731**	0.043	0.6	ug/L	24	Standard
	Ce	140	31461.0	1.2				ug/L	25	Standard
>	**Tb**	159	1048084.0	0.7				ug/L	939701	Standard
	Ho	165	503.3	20.9				ug/L	10	Standard
	Tl	203	28.0	9.4	**0.0013**	0.000	21.0	ug/L	4	Standard
	Tl	205	13.3	21.7	**0.0019**	0.000	19.6	ug/L	2	Standard
	Pb	206	1789.4	1.9	**0.2333**	0.006	2.6	ug/L	240	Standard
	Pb	207	1416.1	2.7	**0.2012**	0.009	4.3	ug/L	198	Standard
	Pb	208	6428.2	1.8	**0.2375**	0.004	1.7	ug/L	883	Standard
	U	238	157.3	5.4	**0.0164**	0.001	7.5	ug/L	3	Standard
>	**Bi**	209	576421.5	0.5				ug/L	523984	Standard

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:12:31
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	65.0	13.3	**0.0052**	0.008	161.5	mg/L	37	Standard
K	39	56.7	10.2	**-0.0092**	0.009	98.8	mg/L	38	Standard
Ca	43	46.7	16.4	**-0.0469**	0.452	963.7	mg/L	10	Standard
Fe	54	615.5	5.2	**0.2993**	0.025	8.2	mg/L	179	Standard
Fe	57	410.0	10.0	**0.0928**	0.096	103.4	mg/L	335	Standard
Sc-1	45	72806.2	1.9				mg/L	61029	Standard
Cl	35	29355.1	3.4				ug/L	20195	Standard
Kr	83	2.3	24.7				ug/L	4	Standard
Br	81	1910.1	15.3				ug/L	1547	Standard
P	31	20493.9	2.3				ug/L	15286	Standard
S	34	2463.5	6.9				ug/L	1572	Standard
Sr	88	83.3	42.1				ug/L	73	Standard
C	12	50.0	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	991.0	1.5				mg/L	6	Standard
Ho-1	165	503.3	20.9				mg/L	10	Standard
Er	166	466.7	4.5				mg/L	13	Standard
I	127	2401.9	23.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		114.911	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		118.217	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:12:31
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.385
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.007
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 19:13:27
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	83999.1	3.6				ug/L	74699	Standard
	Be	9	57157.7	3.6	**57.7443**	0.451	0.8	ug/L	8	Standard
	Al	27	8858742.4	2.0	**53.6133**	0.921	1.7	ug/L	1493	Standard
	Sc	45	70122.0	3.5				ug/L	61029	Standard
	Ti	47	47445.5	0.9	**109.3528**	3.353	3.1	ug/L	35	Standard
	V	51	471838.1	3.0	**54.4025**	0.466	0.9	ug/L	1435	Standard
	Cr	52	466548.5	1.7	**54.0130**	1.261	2.3	ug/L	7635	Standard
	Cr	53	57204.5	1.8	**53.7881**	1.637	3.0	ug/L	363	Standard
	Mn	55	422873.4	1.8	**52.9497**	1.131	2.1	ug/L	747	Standard
	Co	59	420575.5	1.8	**55.6399**	1.134	2.0	ug/L	234	Standard
	Ni	60	114954.2	1.8	**53.5507**	1.121	2.1	ug/L	215	Standard
	Cu	65	113826.7	2.0	**53.3592**	1.065	2.0	ug/L	104	Standard
	Zn	66	66150.5	1.8	**52.4882**	1.217	2.3	ug/L	363	Standard
>	Ge	72	566689.7	3.9				ug/L	495142	Standard
	As	75	66474.1	0.9	**51.9549**	1.668	3.2	ug/L	-50	Standard
	Se	82	6669.2	0.5	**53.0020**	1.808	3.4	ug/L	14	Standard
	Se-1	77	4569.4	1.3	**54.0883**	1.963	3.6	ug/L	42	Standard
>	Ga	71	60.0	25.0				mg/L	42	Standard
	Rb	85	973.4	4.9				ug/L	17	Standard
	Y	89	482055.4	0.8				ug/L	417098	Standard
>	Rh	103	21.7	35.3				ug/L	13	Standard
	Mo	98	311127.6	1.9	**101.8685**	1.475	1.4	ug/L	11	Standard
	Ag	107	381455.1	1.8	**51.0679**	0.625	1.2	ug/L	68	Standard
	Cd	111	122240.5	1.1	**51.0178**	0.254	0.5	mg/L	2	Standard
	Cd	114	297421.0	1.0	**50.6157**	0.803	1.6	ug/L	12	Standard
>	In	115	673075.7	0.6				ug/L	587441	Standard
	Sn	118	341770.5	1.5	**50.9666**	0.480	0.9	ug/L	497	Standard
	Sb	123	266115.5	1.9	**49.7407**	0.724	1.5	ug/L	78	Standard
	Ba	135	114468.8	0.7	**48.7952**	0.078	0.2	ug/L	24	Standard
	Ce	140	298.3	40.9				ug/L	25	Standard
>	Tb	159	1044481.4	1.5				ug/L	939701	Standard
	Ho	165	18.3	83.3				ug/L	10	Standard
	Tl	203	467319.7	1.5	**52.0663**	0.128	0.2	ug/L	4	Standard
	Tl	205	421316.6	1.5	**54.0865**	0.384	0.7	ug/L	2	Standard
	Pb	206	294059.2	2.1	**50.4836**	0.259	0.5	ug/L	240	Standard
	Pb	207	260305.8	2.4	**50.2976**	0.406	0.8	ug/L	198	Standard
	Pb	208	1086866.5	0.8	**53.1659**	0.550	1.0	ug/L	883	Standard
	U	238	395799.2	0.7	**52.8135**	0.760	1.4	ug/L	3	Standard
>	Bi	209	570529.8	1.7				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:15:44
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4777.4	3.9	**5.3091**	0.347	6.5	mg/L	37	Standard
K	39	2250.2	3.7	**4.2173**	0.198	4.7	mg/L	38	Standard
Ca	43	128.3	12.5	**4.5532**	1.080	23.7	mg/L	10	Standard
Fe	54	7377.7	1.8	**5.3086**	0.144	2.7	mg/L	179	Standard
Fe	57	2055.1	1.5	**5.0082**	0.127	2.5	mg/L	335	Standard
> **Sc-1**	45	70122.0	3.5				mg/L	61029	Standard
Cl	35	30001.0	0.9				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	2183.5	6.8				ug/L	1547	Standard
P	31	22995.8	1.5				ug/L	15286	Standard
S	34	2830.3	2.9				ug/L	1572	Standard
Sr	88	66.7	24.1				ug/L	73	Standard
C	12	80.0	50.0				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	9.0	108.0				mg/L	6	Standard
Ho-1	165	18.3	83.3				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	1968.5	24.7				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	115.489		
	Al	27	107.227		
	Sc	45			
	Ti	47	109.353		
	V	51	108.805		
	Cr	52	108.026		
	Cr	53			
	Mn	55	105.899		
	Co	59	111.280		
	Ni	60	107.101		
	Cu	65	106.718		
	Zn	66	104.976		
>	Ge	72		114.450	
	As	75	103.910		
	Se	82	106.004		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:15:44
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.868		
Ag	107	102.136		
Cd	111	102.036		
Cd	114			
In	115		114.578	
Sn	118	101.933		
Sb	123	99.481		
Ba	135	97.590		
Ce	140			
Tb	159			
Ho	165			
Tl	203	104.133		
Tl	205			
Pb	206			
Pb	207			
Pb	208	106.332		
U	238	105.627		
Bi	209		108.883	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:15:44
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 19:16:39
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83692.3	2.8				ug/L	74699	Standard
	Be	9	18.3	15.7	**0.0263**	0.003	12.9	ug/L	8	Standard
	Al	27	3213.7	25.4	**0.0061**	0.004	71.9	ug/L	1493	Standard
	Sc	45	70726.5	3.0				ug/L	61029	Standard
	Ti	47	41.0	12.2	**-0.0158**	0.012	73.1	ug/L	35	Standard
	V	51	1525.1	1.9	**-0.0214**	0.005	22.4	ug/L	1435	Standard
	Cr	52	7659.9	0.9	**-0.1324**	0.015	11.0	ug/L	7635	Standard
	Cr	53	410.0	16.0	**0.0526**	0.061	116.5	ug/L	363	Standard
	Mn	55	997.7	7.3	**-0.0244**	0.010	41.0	ug/L	747	Standard
	Co	59	260.3	1.6	**-0.0033**	0.000	10.0	ug/L	234	Standard
	Ni	60	195.0	3.6	**-0.0341**	0.004	11.2	ug/L	215	Standard
	Cu	65	140.7	10.7	**-0.0385**	0.007	19.0	ug/L	104	Standard
	Zn	66	386.7	3.7	**-0.1493**	0.008	5.7	ug/L	363	Standard
>	**Ge**	72	574185.4	0.9				ug/L	495142	Standard
	As	75	-49.3	11.6	**0.0216**	0.004	19.9	ug/L	-50	Standard
	Se	82	18.5	26.4	**0.0072**	0.039	544.3	ug/L	14	Standard
	Se-1	77	51.3	18.1	**-0.0155**	0.115	739.5	ug/L	42	Standard
>	**Ga**	71	41.7	30.2				mg/L	42	Standard
	Rb	85	33.3	22.9				ug/L	17	Standard
	Y	89	508899.9	2.4				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	287.2	24.5	**0.0857**	0.020	22.8	ug/L	11	Standard
	Ag	107	140.0	24.8	**0.0054**	0.004	76.0	ug/L	68	Standard
	Cd	111	18.8	30.5	**-0.0003**	0.002	828.9	mg/L	2	Standard
	Cd	114	42.6	5.3	**0.0050**	0.000	2.5	ug/L	12	Standard
>	**In**	115	688571.2	3.6				ug/L	587441	Standard
	Sn	118	896.7	13.9	**0.0365**	0.020	55.6	ug/L	497	Standard
	Sb	123	419.7	28.4	**0.0731**	0.025	33.8	ug/L	78	Standard
	Ba	135	58.0	33.3	**0.0081**	0.008	93.7	ug/L	24	Standard
	Ce	140	43.3	17.6				ug/L	25	Standard
>	**Tb**	159	1032903.3	2.1				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	69.3	37.1	**0.0059**	0.003	47.0	ug/L	4	Standard
	Tl	205	78.3	57.2	**0.0101**	0.006	55.2	ug/L	2	Standard
	Pb	206	303.3	6.2	**-0.0195**	0.003	14.1	ug/L	240	Standard
	Pb	207	260.7	8.9	**-0.0200**	0.005	24.1	ug/L	198	Standard
	Pb	208	1157.4	6.9	**-0.0179**	0.003	17.6	ug/L	883	Standard
	U	238	52.0	46.8	**0.0024**	0.003	129.1	ug/L	3	Standard
>	**Bi**	209	575232.2	1.3				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:18:56
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	36.7	28.4	**-0.0241**	0.012	50.4	mg/L	37	Standard
K	39	50.0	26.5	**-0.0190**	0.023	121.8	mg/L	38	Standard
Ca	43	26.7	39.0	**-1.0625**	0.584	55.0	mg/L	10	Standard
Fe	54	147.5	7.0	**-0.0305**	0.005	15.7	mg/L	179	Standard
Fe	57	216.7	8.1	**-0.4394**	0.033	7.6	mg/L	335	Standard
Sc-1	45	70726.5	3.0				mg/L	61029	Standard
Cl	35	29273.6	1.2				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2073.5	10.2				ug/L	1547	Standard
P	31	23289.7	10.5				ug/L	15286	Standard
S	34	2535.2	3.8				ug/L	1572	Standard
Sr	88	68.3	37.5				ug/L	73	Standard
C	12	66.7	22.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	16.3	126.2				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1766.8	31.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.964	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:18:56
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.215
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.780
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:18:56
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138103
Sample Date/Time: Wednesday, June 29, 2016 19:19:52
Number of Replicates: 3
Autosampler Position: 256
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88921.9	2.3				ug/L	74699	Standard
	Be	9	1191.7	0.9	**1.1451**	0.025	2.2	ug/L	8	Standard
	Al	27	19716870.1	2.2	**112.7701**	5.073	4.5	ug/L	1493	Standard
	Sc	45	82006.5	3.2				ug/L	61029	Standard
	Ti	47	8807.2	1.6	**20.0178**	0.269	1.3	ug/L	35	Standard
	V	51	208539.0	1.3	**23.7197**	0.153	0.6	ug/L	1435	Standard
	Cr	52	166877.0	1.1	**18.4795**	0.081	0.4	ug/L	7635	Standard
	Cr	53	19956.5	0.7	**18.3731**	0.126	0.7	ug/L	363	Standard
	Mn	55	4948296.7	1.6	**615.4227**	6.037	1.0	ug/L	747	Standard
	Co	59	82754.8	2.1	**10.8165**	0.154	1.4	ug/L	234	Standard
	Ni	60	51698.7	0.5	**23.7942**	0.298	1.3	ug/L	215	Standard
	Cu	65	48855.0	1.6	**22.6321**	0.340	1.5	ug/L	104	Standard
	Zn	66	105467.2	1.2	**83.1726**	0.384	0.5	ug/L	363	Standard
>	**Ge**	72	571655.8	0.7				ug/L	495142	Standard
	As	75	9636.9	0.9	**7.5116**	0.042	0.6	ug/L	-50	Standard
	Se	82	70.5	8.3	**0.4184**	0.042	10.1	ug/L	14	Standard
	Se-1	77	129.0	9.9	**0.9077**	0.153	16.9	ug/L	42	Standard
>	**Ga**	71	22341.5	2.4				mg/L	42	Standard
	Rb	85	210691.8	1.9				ug/L	17	Standard
	Y	89	776017.6	0.9				ug/L	417098	Standard
>	**Rh**	103	61.7	24.8				ug/L	13	Standard
	Mo	98	1324.4	9.6	**0.4208**	0.045	10.6	ug/L	11	Standard
	Ag	107	918.4	2.4	**0.1080**	0.004	3.7	ug/L	68	Standard
	Cd	111	891.1	4.2	**0.3578**	0.020	5.5	mg/L	2	Standard
	Cd	114	2180.6	6.0	**0.3625**	0.017	4.8	ug/L	12	Standard
>	**In**	115	684819.6	1.2				ug/L	587441	Standard
	Sn	118	1785.1	3.5	**0.1674**	0.006	3.8	ug/L	497	Standard
	Sb	123	354.4	3.1	**0.0609**	0.003	4.6	ug/L	78	Standard
	Ba	135	461342.5	1.4	**193.3434**	2.631	1.4	ug/L	24	Standard
	Ce	140	1368480.3	1.7				ug/L	25	Standard
>	**Tb**	159	1075718.6	2.2				ug/L	939701	Standard
	Ho	165	22026.0	1.7				ug/L	10	Standard
	Tl	203	1135.0	2.0	**0.1247**	0.001	0.6	ug/L	4	Standard
	Tl	205	1093.4	11.6	**0.1404**	0.014	10.2	ug/L	2	Standard
	Pb	206	147882.6	1.3	**25.3549**	0.415	1.6	ug/L	240	Standard
	Pb	207	119862.3	0.6	**23.1268**	0.424	1.8	ug/L	198	Standard
	Pb	208	520445.5	2.0	**25.4242**	0.898	3.5	ug/L	883	Standard
	U	238	4663.1	0.9	**0.6178**	0.006	0.9	ug/L	3	Standard
>	**Bi**	209	570570.9	1.5				ug/L	523984	Standard

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:22:09
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	126.7	27.7	**0.0563**	0.032	56.3	mg/L	37	Standard
K	39	878.4	3.1	**1.3326**	0.091	6.9	mg/L	38	Standard
Ca	43	426.7	4.7	**17.5293**	0.460	2.6	mg/L	10	Standard
Fe	54	21094.2	3.4	**13.1896**	0.865	6.6	mg/L	179	Standard
Fe	57	5826.1	1.4	**13.6721**	0.491	3.6	mg/L	335	Standard
> **Sc-1**	45	82006.5	3.2				mg/L	61029	Standard
Cl	35	28772.6	2.0				ug/L	20195	Standard
Kr	83	3.0	66.7				ug/L	4	Standard
Br	81	3580.4	3.6				ug/L	1547	Standard
P	31	21750.6	1.3				ug/L	15286	Standard
S	34	2703.6	0.6				ug/L	1572	Standard
Sr	88	178.3	13.8				ug/L	73	Standard
C	12	233.3	9.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	96.7	36.3				mg/L	3	Standard
Dy	164	37298.5	1.0				mg/L	6	Standard
Ho-1	165	22026.0	1.7				mg/L	10	Standard
Er	166	20383.7	1.3				mg/L	13	Standard
I	127	19701.1	3.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		119.041	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.453	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:22:09
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	116.577
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.891
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:22:09
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138103
Sample Date/Time: Wednesday, June 29, 2016 19:23:03
Number of Replicates: 3
Autosampler Position: 257
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82351.7	3.3				ug/L	74699	Standard
	Be	9	43.3	24.0	**0.0523**	0.011	20.3	ug/L	8	Standard
	Al	27	430174.7	3.0	**2.6419**	0.016	0.6	ug/L	1493	Standard
	Sc	45	73040.6	1.2				ug/L	61029	Standard
	Ti	47	215.7	6.2	**0.3839**	0.033	8.6	ug/L	35	Standard
	V	51	5937.3	2.7	**0.4853**	0.017	3.5	ug/L	1435	Standard
	Cr	52	11052.0	1.3	**0.2679**	0.017	6.2	ug/L	7635	Standard
	Cr	53	723.4	5.6	**0.3480**	0.038	11.0	ug/L	363	Standard
	Mn	55	109527.8	0.6	**13.4774**	0.102	0.8	ug/L	747	Standard
	Co	59	1998.1	1.8	**0.2248**	0.006	2.5	ug/L	234	Standard
	Ni	60	1283.7	2.7	**0.4700**	0.016	3.4	ug/L	215	Standard
	Cu	65	1177.7	1.9	**0.4444**	0.009	2.1	ug/L	104	Standard
	Zn	66	4258.3	1.3	**2.9218**	0.029	1.0	ug/L	363	Standard
>	**Ge**	72	571680.5	0.6				ug/L	495142	Standard
	As	75	130.4	5.6	**0.1604**	0.006	3.7	ug/L	-50	Standard
	Se	82	14.1	16.1	**-0.0268**	0.018	65.5	ug/L	14	Standard
	Se-1	77	58.0	6.9	**0.0653**	0.044	66.8	ug/L	42	Standard
>	**Ga**	71	521.7	4.0				mg/L	42	Standard
	Rb	85	4554.0	2.0				ug/L	17	Standard
	Y	89	512515.9	2.6				ug/L	417098	Standard
>	**Rh**	103	11.7	24.7				ug/L	13	Standard
	Mo	98	90.4	4.4	**0.0229**	0.001	6.5	ug/L	11	Standard
	Ag	107	144.3	4.6	**0.0058**	0.001	16.1	ug/L	68	Standard
	Cd	111	21.8	34.6	**0.0009**	0.003	335.6	mg/L	2	Standard
	Cd	114	67.8	46.0	**0.0090**	0.005	55.6	ug/L	12	Standard
>	**In**	115	696556.9	1.5				ug/L	587441	Standard
	Sn	118	640.0	3.6	**-0.0023**	0.005	204.9	ug/L	497	Standard
	Sb	123	182.0	15.1	**0.0287**	0.006	19.2	ug/L	78	Standard
	Ba	135	9992.3	2.7	**4.1009**	0.069	1.7	ug/L	24	Standard
	Ce	140	29587.2	2.0				ug/L	25	Standard
>	**Tb**	159	1058949.6	2.5				ug/L	939701	Standard
	Ho	165	493.3	6.5				ug/L	10	Standard
	Tl	203	41.7	21.8	**0.0028**	0.001	37.1	ug/L	4	Standard
	Tl	205	43.3	54.5	**0.0057**	0.003	52.4	ug/L	2	Standard
	Pb	206	3583.8	1.2	**0.5358**	0.012	2.2	ug/L	240	Standard
	Pb	207	2855.9	1.4	**0.4744**	0.011	2.4	ug/L	198	Standard
	Pb	208	12510.0	0.4	**0.5295**	0.006	1.1	ug/L	883	Standard
	U	238	110.3	6.2	**0.0101**	0.001	8.0	ug/L	3	Standard
>	**Bi**	209	579224.4	1.1				ug/L	523984	Standard

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:25:20
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	51.7	29.6	**-0.0092**	0.017	184.4	mg/L	37	Standard
K	39	41.7	61.6	**-0.0375**	0.047	124.3	mg/L	38	Standard
Ca	43	26.7	10.8	**-1.1162**	0.146	13.1	mg/L	10	Standard
Fe	54	584.0	15.6	**0.2752**	0.062	22.7	mg/L	179	Standard
Fe	57	388.3	5.2	**0.0289**	0.063	219.5	mg/L	335	Standard
> **Sc-1**	45	73040.6	1.2				mg/L	61029	Standard
Cl	35	28933.6	1.8				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2263.5	5.9				ug/L	1547	Standard
P	31	20116.7	3.4				ug/L	15286	Standard
S	34	2530.2	7.0				ug/L	1572	Standard
Sr	88	75.0	23.1				ug/L	73	Standard
C	12	66.7	56.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	963.7	4.1				mg/L	6	Standard
Ho-1	165	493.3	6.5				mg/L	10	Standard
Er	166	410.0	6.5				mg/L	13	Standard
I	127	2373.5	31.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.245	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.458	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:25:20
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		118.575
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.542
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138104
Sample Date/Time: Wednesday, June 29, 2016 19:26:15
Number of Replicates: 3
Autosampler Position: 258
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85261.0	2.3				ug/L	74699	Standard
	Be	9	1065.0	16.2	**1.0661**	0.158	14.8	ug/L	8	Standard
	Al	27	15142099.3	1.2	**90.2750**	1.228	1.4	ug/L	1493	Standard
	Sc	45	77933.3	2.2				ug/L	61029	Standard
	Ti	47	7118.0	1.5	**16.2707**	0.118	0.7	ug/L	35	Standard
	V	51	161044.2	0.8	**18.4034**	0.196	1.1	ug/L	1435	Standard
	Cr	52	146045.2	0.2	**16.1669**	0.325	2.0	ug/L	7635	Standard
	Cr	53	17793.8	1.2	**16.4664**	0.462	2.8	ug/L	363	Standard
	Mn	55	4299510.8	1.0	**538.5113**	8.104	1.5	ug/L	747	Standard
	Co	59	67815.8	1.0	**8.9211**	0.182	2.0	ug/L	234	Standard
	Ni	60	46536.4	1.1	**21.5551**	0.152	0.7	ug/L	215	Standard
	Cu	65	28919.6	1.2	**13.4506**	0.294	2.2	ug/L	104	Standard
	Zn	66	52781.1	1.6	**41.6917**	0.739	1.8	ug/L	363	Standard
>	**Ge**	72	567729.5	1.7				ug/L	495142	Standard
	As	75	4550.5	2.3	**3.6043**	0.141	3.9	ug/L	-50	Standard
	Se	82	43.5	41.5	**0.2091**	0.148	70.8	ug/L	14	Standard
	Se-1	77	118.3	4.0	**0.7906**	0.045	5.7	ug/L	42	Standard
>	**Ga**	71	21335.0	2.6				mg/L	42	Standard
	Rb	85	199689.0	1.1				ug/L	17	Standard
	Y	89	717025.3	1.7				ug/L	417098	Standard
>	**Rh**	103	53.3	39.0				ug/L	13	Standard
	Mo	98	750.2	3.3	**0.2455**	0.003	1.3	ug/L	11	Standard
	Ag	107	572.7	8.1	**0.0655**	0.005	7.5	ug/L	68	Standard
	Cd	111	313.3	5.6	**0.1259**	0.007	5.6	mg/L	2	Standard
	Cd	114	729.6	4.4	**0.1250**	0.007	5.3	ug/L	12	Standard
>	**In**	115	657920.1	2.1				ug/L	587441	Standard
	Sn	118	513.3	6.6	**-0.0162**	0.005	32.2	ug/L	497	Standard
	Sb	123	156.7	10.3	**0.0258**	0.004	14.3	ug/L	78	Standard
	Ba	135	278778.0	0.4	**121.6374**	2.867	2.4	ug/L	24	Standard
	Ce	140	1205012.6	0.9				ug/L	25	Standard
>	**Tb**	159	1028655.6	1.0				ug/L	939701	Standard
	Ho	165	19245.5	0.5				ug/L	10	Standard
	Tl	203	1022.7	4.2	**0.1168**	0.004	3.6	ug/L	4	Standard
	Tl	205	930.0	6.2	**0.1244**	0.008	6.1	ug/L	2	Standard
	Pb	206	75172.7	1.1	**13.3810**	0.239	1.8	ug/L	240	Standard
	Pb	207	60753.2	0.6	**12.1662**	0.152	1.3	ug/L	198	Standard
	Pb	208	268421.8	0.9	**13.6089**	0.053	0.4	ug/L	883	Standard
	U	238	4956.8	1.6	**0.6840**	0.014	2.1	ug/L	3	Standard
>	**Bi**	209	548179.9	0.7				ug/L	523984	Standard

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:28:31
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**29.7760**	51.565	173.2	mg/L	0	Standard
Mg	24	116.7	8.9	**0.0527**	0.008	15.2	mg/L	37	Standard
K	39	606.7	4.1	**0.9368**	0.065	7.0	mg/L	38	Standard
Ca	43	223.3	26.5	**8.5548**	3.074	35.9	mg/L	10	Standard
Fe	54	18758.0	3.4	**12.3136**	0.167	1.4	mg/L	179	Standard
Fe	57	5059.2	5.4	**12.3881**	0.534	4.3	mg/L	335	Standard
Sc-1	45	77933.3	2.2				mg/L	61029	Standard
Cl	35	27511.6	2.5				ug/L	20195	Standard
Kr	83	4.0	50.0				ug/L	4	Standard
Br	81	2450.2	6.4				ug/L	1547	Standard
P	31	19838.0	2.4				ug/L	15286	Standard
S	34	2438.5	4.0				ug/L	1572	Standard
Sr	88	130.0	10.2				ug/L	73	Standard
C	12	76.7	41.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	16.7	124.9				mg/L	3	Standard
Dy	164	31319.3	1.6				mg/L	6	Standard
Ho-1	165	19245.5	0.5				mg/L	10	Standard
Er	166	17009.5	2.0				mg/L	13	Standard
I	127	14255.0	0.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		114.140	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.660	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:28:31
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.998
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.618
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138104
Sample Date/Time: Wednesday, June 29, 2016 19:29:26
Number of Replicates: 3
Autosampler Position: 259
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81019.4	0.7				ug/L	74699	Standard
	Be	9	41.7	18.3	**0.0512**	0.008	15.3	ug/L	8	Standard
	Al	27	329954.8	1.9	**2.0568**	0.044	2.2	ug/L	1493	Standard
	Sc	45	70853.8	4.4				ug/L	61029	Standard
	Ti	47	170.0	4.8	**0.2650**	0.022	8.2	ug/L	35	Standard
	V	51	4919.9	1.5	**0.3475**	0.009	2.6	ug/L	1435	Standard
	Cr	52	11144.8	0.5	**0.2299**	0.010	4.3	ug/L	7635	Standard
	Cr	53	798.4	9.1	**0.3907**	0.073	18.6	ug/L	363	Standard
	Mn	55	93006.7	2.2	**10.9903**	0.351	3.2	ug/L	747	Standard
	Co	59	1669.1	4.9	**0.1734**	0.010	5.5	ug/L	234	Standard
	Ni	60	1239.4	3.4	**0.4280**	0.019	4.5	ug/L	215	Standard
	Cu	65	783.4	0.7	**0.2472**	0.006	2.3	ug/L	104	Standard
	Zn	66	3223.7	4.0	**2.0065**	0.120	6.0	ug/L	363	Standard
>	**Ge**	72	593953.7	1.0				ug/L	495142	Standard
	As	75	25.6	116.2	**0.0786**	0.022	28.0	ug/L	-50	Standard
	Se	82	15.5	11.3	**-0.0201**	0.014	71.9	ug/L	14	Standard
	Se-1	77	55.7	3.7	**0.0130**	0.019	145.3	ug/L	42	Standard
>	**Ga**	71	496.7	9.6				mg/L	42	Standard
	Rb	85	4318.9	1.6				ug/L	17	Standard
	Y	89	493099.4	1.2				ug/L	417098	Standard
>	**Rh**	103	8.3	91.7				ug/L	13	Standard
	Mo	98	63.8	10.3	**0.0154**	0.002	13.8	ug/L	11	Standard
	Ag	107	130.3	15.7	**0.0047**	0.003	56.8	ug/L	68	Standard
	Cd	111	7.2	21.0	**-0.0049**	0.001	12.8	mg/L	2	Standard
	Cd	114	27.0	44.7	**0.0025**	0.002	80.8	ug/L	12	Standard
>	**In**	115	665068.2	0.5				ug/L	587441	Standard
	Sn	118	545.0	15.7	**-0.0123**	0.013	107.4	ug/L	497	Standard
	Sb	123	150.6	9.1	**0.0243**	0.003	11.0	ug/L	78	Standard
	Ba	135	6015.5	0.5	**2.5800**	0.016	0.6	ug/L	24	Standard
	Ce	140	25678.4	3.6				ug/L	25	Standard
>	**Tb**	159	1023204.1	0.9				ug/L	939701	Standard
	Ho	165	420.0	5.5				ug/L	10	Standard
	Tl	203	37.3	22.8	**0.0024**	0.001	40.1	ug/L	4	Standard
	Tl	205	30.0	50.0	**0.0040**	0.002	46.9	ug/L	2	Standard
	Pb	206	1995.8	4.0	**0.2734**	0.016	6.0	ug/L	240	Standard
	Pb	207	1576.4	5.4	**0.2364**	0.019	8.0	ug/L	198	Standard
	Pb	208	6805.9	1.7	**0.2606**	0.003	1.1	ug/L	883	Standard
	U	238	99.3	2.5	**0.0089**	0.000	2.4	ug/L	3	Standard
>	**Bi**	209	568219.3	0.9				ug/L	523984	Standard

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:31:43
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	30.0	66.7	**-0.0312**	0.023	74.2	mg/L	37	Standard
K	39	65.0	30.8	**0.0108**	0.043	395.4	mg/L	38	Standard
Ca	43	20.0	0.0	**-1.4338**	0.047	3.3	mg/L	10	Standard
Fe	54	499.3	4.0	**0.2265**	0.010	4.2	mg/L	179	Standard
Fe	57	346.7	6.8	**-0.0592**	0.056	94.8	mg/L	335	Standard
> **Sc-1**	45	70853.8	4.4				mg/L	61029	Standard
Cl	35	26917.9	3.4				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	1996.8	2.8				ug/L	1547	Standard
P	31	19530.9	2.5				ug/L	15286	Standard
S	34	2235.2	2.4				ug/L	1572	Standard
Sr	88	70.0	18.9				ug/L	73	Standard
C	12	83.3	38.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	818.0	9.1				mg/L	6	Standard
Ho-1	165	420.0	5.5				mg/L	10	Standard
Er	166	390.0	11.8				mg/L	13	Standard
I	127	1930.1	32.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		108.461	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		119.956	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:31:43
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.214
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.442
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138105
Sample Date/Time: Wednesday, June 29, 2016 19:32:37
Number of Replicates: 3
Autosampler Position: 260
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86630.3	1.6				ug/L	74699	Standard
	Be	9	1345.1	15.8	**1.3230**	0.188	14.2	ug/L	8	Standard
	Al	27	25518871.0	0.5	**149.7364**	2.025	1.4	ug/L	1493	Standard
	Sc	45	82793.9	1.1				ug/L	61029	Standard
	Ti	47	9047.0	1.1	**21.0343**	0.216	1.0	ug/L	35	Standard
	V	51	241144.8	2.1	**28.0828**	0.136	0.5	ug/L	1435	Standard
	Cr	52	185820.6	3.2	**21.1812**	0.417	2.0	ug/L	7635	Standard
	Cr	53	22667.0	3.7	**21.3886**	0.515	2.4	ug/L	363	Standard
	Mn	55	5666490.2	1.7	**720.7034**	2.244	0.3	ug/L	747	Standard
	Co	59	89984.1	1.6	**12.0322**	0.097	0.8	ug/L	234	Standard
	Ni	60	57928.9	2.2	**27.2782**	0.178	0.7	ug/L	215	Standard
	Cu	65	41127.1	3.0	**19.4650**	0.281	1.4	ug/L	104	Standard
	Zn	66	70831.6	3.1	**56.9699**	0.880	1.5	ug/L	363	Standard
>	**Ge**	72	559032.2	1.6				ug/L	495142	Standard
	As	75	9088.7	3.2	**7.2452**	0.123	1.7	ug/L	-50	Standard
	Se	82	66.3	23.6	**0.3960**	0.117	29.7	ug/L	14	Standard
	Se-1	77	130.0	4.3	**0.9538**	0.045	4.7	ug/L	42	Standard
>	**Ga**	71	25715.1	2.4				mg/L	42	Standard
	Rb	85	243499.9	2.1				ug/L	17	Standard
	Y	89	766854.2	0.5				ug/L	417098	Standard
>	**Rh**	103	51.7	5.6				ug/L	13	Standard
	Mo	98	760.8	5.6	**0.2496**	0.019	7.7	ug/L	11	Standard
	Ag	107	727.0	8.3	**0.0869**	0.010	11.3	ug/L	68	Standard
	Cd	111	420.9	5.2	**0.1721**	0.013	7.6	mg/L	2	Standard
	Cd	114	1021.1	15.6	**0.1761**	0.030	17.3	ug/L	12	Standard
>	**In**	115	657583.1	2.2				ug/L	587441	Standard
	Sn	118	706.7	9.4	**0.0135**	0.012	91.4	ug/L	497	Standard
	Sb	123	160.0	3.9	**0.0265**	0.002	7.0	ug/L	78	Standard
	Ba	135	489546.4	1.0	**213.7510**	6.703	3.1	ug/L	24	Standard
	Ce	140	1473172.3	0.9				ug/L	25	Standard
>	**Tb**	159	1043758.3	1.4				ug/L	939701	Standard
	Ho	165	23187.7	0.9				ug/L	10	Standard
	Tl	203	1287.4	7.3	**0.1458**	0.012	8.5	ug/L	4	Standard
	Tl	205	1168.4	5.5	**0.1545**	0.010	6.8	ug/L	2	Standard
	Pb	206	110607.4	2.3	**19.4839**	0.695	3.6	ug/L	240	Standard
	Pb	207	89635.8	2.2	**17.7667**	0.631	3.6	ug/L	198	Standard
	Pb	208	392577.0	0.9	**19.6965**	0.434	2.2	ug/L	883	Standard
	U	238	8119.8	0.8	**1.1095**	0.020	1.8	ug/L	3	Standard
>	**Bi**	209	554986.0	1.4				ug/L	523984	Standard

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:34:54
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**42.5482**	42.299	99.4	mg/L	0	Standard
Mg	24	165.0	16.0	**0.0921**	0.027	29.3	mg/L	37	Standard
K	39	870.0	2.5	**1.3032**	0.021	1.6	mg/L	38	Standard
Ca	43	623.3	6.5	**26.5121**	2.122	8.0	mg/L	10	Standard
Fe	54	24836.5	3.4	**15.3836**	0.500	3.3	mg/L	179	Standard
Fe	57	6766.5	3.6	**15.8771**	0.513	3.2	mg/L	335	Standard
Sc-1	45	82793.9	1.1				mg/L	61029	Standard
Cl	35	27591.1	0.8				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2510.2	7.0				ug/L	1547	Standard
P	31	20592.3	2.0				ug/L	15286	Standard
S	34	2640.2	1.7				ug/L	1572	Standard
Sr	88	220.0	6.0				ug/L	73	Standard
C	12	86.7	24.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	46.7	44.6				mg/L	3	Standard
Dy	164	40115.6	1.1				mg/L	6	Standard
Ho-1	165	23187.7	0.9				mg/L	10	Standard
Er	166	21645.5	1.0				mg/L	13	Standard
I	127	13629.5	5.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		115.973	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		112.903	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:34:54
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.940
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.916
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:34:54
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138105
Sample Date/Time: Wednesday, June 29, 2016 19:35:50
Number of Replicates: 3
Autosampler Position: 401
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81775.3	3.6				ug/L	74699	Standard
	Be	9	45.0	11.1	**0.0544**	0.006	11.5	ug/L	8	Standard
	Al	27	547945.2	2.8	**3.3933**	0.048	1.4	ug/L	1493	Standard
	Sc	45	73444.3	1.5				ug/L	61029	Standard
	Ti	47	212.7	8.9	**0.3795**	0.037	9.9	ug/L	35	Standard
	V	51	6482.3	2.1	**0.5523**	0.015	2.6	ug/L	1435	Standard
	Cr	52	11556.7	2.9	**0.3347**	0.014	4.0	ug/L	7635	Standard
	Cr	53	865.0	10.6	**0.4848**	0.072	14.9	ug/L	363	Standard
	Mn	55	123099.8	1.9	**15.2575**	0.218	1.4	ug/L	747	Standard
	Co	59	2101.5	1.5	**0.2401**	0.008	3.5	ug/L	234	Standard
	Ni	60	1463.7	4.5	**0.5572**	0.024	4.3	ug/L	215	Standard
	Cu	65	1004.0	2.5	**0.3664**	0.012	3.3	ug/L	104	Standard
	Zn	66	6132.9	2.2	**4.4370**	0.037	0.8	ug/L	363	Standard
>	**Ge**	72	568314.4	1.9				ug/L	495142	Standard
	As	75	135.4	21.1	**0.1651**	0.024	14.5	ug/L	-50	Standard
	Se	82	20.1	25.4	**0.0217**	0.043	197.6	ug/L	14	Standard
	Se-1	77	47.3	2.4	**-0.0575**	0.019	32.3	ug/L	42	Standard
>	**Ga**	71	615.0	5.6				mg/L	42	Standard
	Rb	85	5134.2	2.0				ug/L	17	Standard
	Y	89	506159.2	1.9				ug/L	417098	Standard
>	**Rh**	103	11.7	107.9				ug/L	13	Standard
	Mo	98	50.7	18.0	**0.0105**	0.003	28.3	ug/L	11	Standard
	Ag	107	125.7	10.6	**0.0035**	0.002	48.0	ug/L	68	Standard
	Cd	111	10.9	48.5	**-0.0035**	0.002	62.7	mg/L	2	Standard
	Cd	114	42.4	34.4	**0.0049**	0.002	48.6	ug/L	12	Standard
>	**In**	115	689821.8	0.4				ug/L	587441	Standard
	Sn	118	595.0	5.8	**-0.0080**	0.005	59.8	ug/L	497	Standard
	Sb	123	118.9	14.7	**0.0175**	0.003	18.8	ug/L	78	Standard
	Ba	135	10423.6	1.2	**4.3208**	0.035	0.8	ug/L	24	Standard
	Ce	140	31651.4	0.9				ug/L	25	Standard
>	**Tb**	159	1043112.5	1.4				ug/L	939701	Standard
	Ho	165	500.0	7.2				ug/L	10	Standard
	Tl	203	34.0	12.8	**0.0020**	0.000	23.2	ug/L	4	Standard
	Tl	205	20.0	66.1	**0.0027**	0.002	61.7	ug/L	2	Standard
	Pb	206	2645.2	1.4	**0.3784**	0.005	1.3	ug/L	240	Standard
	Pb	207	2160.5	2.6	**0.3433**	0.011	3.1	ug/L	198	Standard
	Pb	208	9496.8	0.4	**0.3858**	0.004	1.1	ug/L	883	Standard
	U	238	183.7	2.7	**0.0198**	0.001	4.5	ug/L	3	Standard
>	**Bi**	209	577076.9	1.0				ug/L	523984	Standard

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:38:07
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**15.9398**	27.600	173.2	mg/L	0	Standard
Mg	24	31.7	39.7	**-0.0310**	0.014	45.0	mg/L	37	Standard
K	39	75.0	17.6	**0.0236**	0.024	101.0	mg/L	38	Standard
Ca	43	45.0	44.4	**-0.1502**	1.087	724.0	mg/L	10	Standard
Fe	54	679.1	0.8	**0.3402**	0.010	2.9	mg/L	179	Standard
Fe	57	390.0	9.7	**0.0271**	0.102	376.6	mg/L	335	Standard
Sc-1	45	73444.3	1.5				mg/L	61029	Standard
Cl	35	27695.3	0.4				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	2046.8	15.0				ug/L	1547	Standard
P	31	20195.1	3.7				ug/L	15286	Standard
S	34	2511.9	7.3				ug/L	1572	Standard
Sr	88	68.3	42.9				ug/L	73	Standard
C	12	50.0	40.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	1042.0	9.5				mg/L	6	Standard
Ho-1	165	500.0	7.2				mg/L	10	Standard
Er	166	446.7	15.2				mg/L	13	Standard
I	127	1841.8	32.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		109.473	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.778	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:38:07
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.428
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.132
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138107
Sample Date/Time: Wednesday, June 29, 2016 19:39:01
Number of Replicates: 3
Autosampler Position: 402
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84183.2	0.7				ug/L	74699	Standard
	Be	9	1125.0	5.6	**1.1418**	0.071	6.2	ug/L	8	Standard
	Al	27	16362118.1	1.0	**98.7783**	0.357	0.4	ug/L	1493	Standard
	Sc	45	80079.5	2.6				ug/L	61029	Standard
	Ti	47	5684.7	1.2	**13.0248**	0.178	1.4	ug/L	35	Standard
	V	51	193164.5	0.6	**22.2022**	0.395	1.8	ug/L	1435	Standard
	Cr	52	144926.2	1.3	**16.0993**	0.107	0.7	ug/L	7635	Standard
	Cr	53	17089.6	2.1	**15.8598**	0.165	1.0	ug/L	363	Standard
	Mn	55	4945150.4	1.2	**621.8545**	12.591	2.0	ug/L	747	Standard
	Co	59	79442.1	0.4	**10.4972**	0.101	1.0	ug/L	234	Standard
	Ni	60	48759.0	0.6	**22.6802**	0.156	0.7	ug/L	215	Standard
	Cu	65	44548.4	1.0	**20.8546**	0.085	0.4	ug/L	104	Standard
	Zn	66	116987.3	1.0	**93.3289**	1.249	1.3	ug/L	363	Standard
>	**Ge**	72	565487.8	1.3				ug/L	495142	Standard
	As	75	12786.6	1.5	**10.0557**	0.155	1.5	ug/L	-50	Standard
	Se	82	93.9	6.5	**0.6112**	0.058	9.6	ug/L	14	Standard
	Se-1	77	134.7	12.3	**0.9941**	0.220	22.1	ug/L	42	Standard
>	**Ga**	71	20750.9	1.1				mg/L	42	Standard
	Rb	85	229576.4	1.2				ug/L	17	Standard
	Y	89	760564.3	0.9				ug/L	417098	Standard
>	**Rh**	103	26.7	39.0				ug/L	13	Standard
	Mo	98	714.7	6.9	**0.2279**	0.017	7.3	ug/L	11	Standard
	Ag	107	706.7	2.8	**0.0816**	0.003	4.3	ug/L	68	Standard
	Cd	111	991.8	3.3	**0.4053**	0.015	3.7	mg/L	2	Standard
	Cd	114	2256.1	2.3	**0.3810**	0.005	1.2	ug/L	12	Standard
>	**In**	115	674434.0	1.0				ug/L	587441	Standard
	Sn	118	836.7	10.4	**0.0300**	0.012	41.1	ug/L	497	Standard
	Sb	123	542.0	6.0	**0.0969**	0.005	5.6	ug/L	78	Standard
	Ba	135	363437.5	1.5	**154.6607**	2.916	1.9	ug/L	24	Standard
	Ce	140	1323769.8	0.4				ug/L	25	Standard
>	**Tb**	159	1046974.9	0.8				ug/L	939701	Standard
	Ho	165	21420.1	0.8				ug/L	10	Standard
	Tl	203	1015.7	4.2	**0.1138**	0.005	4.5	ug/L	4	Standard
	Tl	205	906.7	2.2	**0.1191**	0.002	2.0	ug/L	2	Standard
	Pb	206	225485.9	0.9	**39.5300**	0.272	0.7	ug/L	240	Standard
	Pb	207	184080.5	0.2	**36.3178**	0.069	0.2	ug/L	198	Standard
	Pb	208	811320.1	0.8	**40.5199**	0.434	1.1	ug/L	883	Standard
	U	238	5280.9	2.0	**0.7154**	0.017	2.3	ug/L	3	Standard
>	**Bi**	209	558508.9	0.3				ug/L	523984	Standard

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:41:18
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	105.0	49.7	**0.0390**	0.054	139.2	mg/L	37	Standard
K	39	875.0	3.0	**1.3597**	0.017	1.3	mg/L	38	Standard
Ca	43	123.3	10.2	**3.4107**	0.543	15.9	mg/L	10	Standard
Fe	54	18540.3	2.4	**11.8486**	0.489	4.1	mg/L	179	Standard
Fe	57	5050.8	2.1	**12.0198**	0.614	5.1	mg/L	335	Standard
> **Sc-1**	45	80079.5	2.6				mg/L	61029	Standard
Cl	35	26211.3	2.3				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	4010.5	1.9				ug/L	1547	Standard
P	31	18821.7	2.8				ug/L	15286	Standard
S	34	2180.2	5.1				ug/L	1572	Standard
Sr	88	100.0	25.0				ug/L	73	Standard
C	12	166.7	30.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	70.0	28.6				mg/L	3	Standard
Dy	164	35314.8	2.4				mg/L	6	Standard
Ho-1	165	21420.1	0.8				mg/L	10	Standard
Er	166	18965.2	4.5				mg/L	13	Standard
I	127	21211.5	3.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		112.697	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.207	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:41:18
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.809
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.589
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:41:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138107
Sample Date/Time: Wednesday, June 29, 2016 19:42:13
Number of Replicates: 3
Autosampler Position: 403
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79962.1	1.1				ug/L	74699	Standard
	Be	9	35.0	37.8	**0.0449**	0.015	32.4	ug/L	8	Standard
	Al	27	348986.4	2.3	**2.2050**	0.043	1.9	ug/L	1493	Standard
	Sc	45	69050.4	2.1				ug/L	61029	Standard
	Ti	47	147.7	4.0	**0.2190**	0.009	4.2	ug/L	35	Standard
	V	51	5537.0	5.1	**0.4222**	0.040	9.5	ug/L	1435	Standard
	Cr	52	11219.2	0.9	**0.2513**	0.015	5.8	ug/L	7635	Standard
	Cr	53	801.7	3.8	**0.4006**	0.025	6.2	ug/L	363	Standard
	Mn	55	101661.8	3.0	**12.1540**	0.524	4.3	ug/L	747	Standard
	Co	59	1813.1	5.5	**0.1941**	0.015	7.9	ug/L	234	Standard
	Ni	60	1251.7	0.2	**0.4392**	0.009	2.1	ug/L	215	Standard
	Cu	65	1098.7	4.3	**0.3937**	0.026	6.7	ug/L	104	Standard
	Zn	66	4663.4	1.9	**3.1415**	0.081	2.6	ug/L	363	Standard
>	**Ge**	72	587910.1	1.4				ug/L	495142	Standard
	As	75	208.3	5.8	**0.2162**	0.008	3.9	ug/L	-50	Standard
	Se	82	14.6	52.1	**-0.0264**	0.056	213.9	ug/L	14	Standard
	Se-1	77	58.3	15.6	**0.0509**	0.111	217.0	ug/L	42	Standard
>	**Ga**	71	483.3	11.4				mg/L	42	Standard
	Rb	85	4672.4	3.2				ug/L	17	Standard
	Y	89	494582.5	2.5				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	50.9	2.1	**0.0113**	0.001	5.2	ug/L	11	Standard
	Ag	107	137.0	5.3	**0.0059**	0.001	14.4	ug/L	68	Standard
	Cd	111	22.2	22.2	**0.0016**	0.002	123.1	mg/L	2	Standard
	Cd	114	55.9	25.6	**0.0077**	0.003	34.0	ug/L	12	Standard
>	**In**	115	657881.3	1.7				ug/L	587441	Standard
	Sn	118	586.7	15.6	**-0.0052**	0.012	239.2	ug/L	497	Standard
	Sb	123	130.5	4.0	**0.0208**	0.001	5.8	ug/L	78	Standard
	Ba	135	7584.6	1.4	**3.2933**	0.054	1.6	ug/L	24	Standard
	Ce	140	27601.8	1.0				ug/L	25	Standard
>	**Tb**	159	1004279.0	0.4				ug/L	939701	Standard
	Ho	165	470.0	12.8				ug/L	10	Standard
	Tl	203	31.3	21.7	**0.0018**	0.001	42.5	ug/L	4	Standard
	Tl	205	38.3	39.8	**0.0052**	0.002	37.2	ug/L	2	Standard
	Pb	206	5039.5	2.9	**0.8074**	0.036	4.5	ug/L	240	Standard
	Pb	207	4082.6	3.5	**0.7312**	0.039	5.3	ug/L	198	Standard
	Pb	208	17367.6	1.6	**0.7884**	0.025	3.2	ug/L	883	Standard
	U	238	118.3	14.2	**0.0116**	0.002	21.2	ug/L	3	Standard
>	**Bi**	209	562827.5	1.3				ug/L	523984	Standard

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:44:30
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	41.7	6.9	**-0.0175**	0.004	24.7	mg/L	37	Standard
K	39	65.0	46.8	**0.0126**	0.058	464.1	mg/L	38	Standard
Ca	43	20.0	66.1	**-1.4118**	0.719	50.9	mg/L	10	Standard
Fe	54	548.6	17.9	**0.2728**	0.072	26.3	mg/L	179	Standard
Fe	57	318.3	8.7	**-0.1165**	0.103	88.6	mg/L	335	Standard
> **Sc-1**	45	69050.4	2.1				mg/L	61029	Standard
Cl	35	24683.4	2.1				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	2036.8	6.4				ug/L	1547	Standard
P	31	17868.9	0.9				ug/L	15286	Standard
S	34	1880.1	5.1				ug/L	1572	Standard
Sr	88	70.0	32.7				ug/L	73	Standard
C	12	63.3	32.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	851.3	21.1				mg/L	6	Standard
Ho-1	165	470.0	12.8				mg/L	10	Standard
Er	166	460.0	19.0				mg/L	13	Standard
I	127	2083.5	32.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		107.046	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		118.736	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:44:30
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.991
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.413
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138402
Sample Date/Time: Wednesday, June 29, 2016 19:45:24
Number of Replicates: 3
Autosampler Position: 404
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78829.6	2.0				ug/L	74699	Standard
	Be	9	476.7	4.8	**0.5211**	0.034	6.6	ug/L	8	Standard
	Al	27	790696.4	0.9	**5.0858**	0.055	1.1	ug/L	1493	Standard
	Sc	45	79546.6	1.7				ug/L	61029	Standard
	Ti	47	5657.7	0.8	**12.8152**	0.349	2.7	ug/L	35	Standard
	V	51	106854.5	1.6	**12.0511**	0.193	1.6	ug/L	1435	Standard
	Cr	52	263051.4	1.4	**29.7045**	0.729	2.5	ug/L	7635	Standard
	Cr	53	32229.3	4.0	**29.8443**	0.696	2.3	ug/L	363	Standard
	Mn	55	1259215.6	0.6	**156.4341**	3.728	2.4	ug/L	747	Standard
	Co	59	45647.0	0.6	**5.9475**	0.157	2.6	ug/L	234	Standard
	Ni	60	134444.3	1.6	**62.0317**	0.857	1.4	ug/L	215	Standard
	Cu	65	50137.5	0.5	**23.2180**	0.523	2.3	ug/L	104	Standard
	Zn	66	191528.7	0.9	**151.3328**	2.746	1.8	ug/L	363	Standard
>	**Ge**	72	572152.6	2.7				ug/L	495142	Standard
	As	75	5700.3	1.6	**4.4645**	0.048	1.1	ug/L	-50	Standard
	Se	82	330.0	4.2	**2.4632**	0.049	2.0	ug/L	14	Standard
	Se-1	77	314.7	3.0	**3.1071**	0.115	3.7	ug/L	42	Standard
>	**Ga**	71	3685.4	1.6				mg/L	42	Standard
	Rb	85	41053.9	1.7				ug/L	17	Standard
	Y	89	1388727.0	1.6				ug/L	417098	Standard
>	**Rh**	103	25.0	0.0				ug/L	13	Standard
	Mo	98	3897.2	2.0	**1.2782**	0.019	1.5	ug/L	11	Standard
	Ag	107	551.3	5.8	**0.0614**	0.003	5.7	ug/L	68	Standard
	Cd	111	1240.5	0.7	**0.5131**	0.007	1.4	mg/L	2	Standard
	Cd	114	2915.5	4.4	**0.4970**	0.020	4.1	ug/L	12	Standard
>	**In**	115	669008.5	1.2				ug/L	587441	Standard
	Sn	118	676.7	8.2	**0.0070**	0.008	114.7	ug/L	497	Standard
	Sb	123	186.6	3.8	**0.0309**	0.001	3.8	ug/L	78	Standard
	Ba	135	221251.7	0.3	**94.9139**	1.345	1.4	ug/L	24	Standard
	Ce	140	471655.7	0.9				ug/L	25	Standard
>	**Tb**	159	1067869.0	0.9				ug/L	939701	Standard
	Ho	165	47698.3	1.5				ug/L	10	Standard
	Tl	203	1397.7	2.5	**0.1543**	0.005	3.4	ug/L	4	Standard
	Tl	205	1176.7	6.0	**0.1515**	0.008	5.0	ug/L	2	Standard
	Pb	206	21429.5	1.2	**3.6210**	0.069	1.9	ug/L	240	Standard
	Pb	207	15728.2	0.6	**2.9798**	0.021	0.7	ug/L	198	Standard
	Pb	208	70387.8	1.2	**3.3811**	0.084	2.5	ug/L	883	Standard
	U	238	18087.1	1.2	**2.4141**	0.054	2.2	ug/L	3	Standard
>	**Bi**	209	569388.1	1.3				ug/L	523984	Standard

Sample ID: L1606138402
Report Date/Time: Wednesday, June 29, 2016 19:47:41
Page 1

Approved: July 04, 2016

Na	23	5.0	173.2	**43.9188**	76.061	173.2	mg/L	0	Standard
Mg	24	210.0	6.3	**0.1429**	0.013	9.0	mg/L	37	Standard
K	39	275.0	8.3	**0.3521**	0.037	10.4	mg/L	38	Standard
Ca	43	148.3	19.8	**4.6742**	1.471	31.5	mg/L	10	Standard
Fe	54	12478.7	2.5	**7.9817**	0.313	3.9	mg/L	179	Standard
Fe	57	3430.4	4.1	**7.8688**	0.221	2.8	mg/L	335	Standard
> **Sc-1**	45	79546.6	1.7				mg/L	61029	Standard
Cl	35	26111.8	0.7				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	2043.5	3.5				ug/L	1547	Standard
P	31	19195.5	2.0				ug/L	15286	Standard
S	34	2368.5	5.4				ug/L	1572	Standard
Sr	88	153.3	13.2				ug/L	73	Standard
C	12	70.0	14.3				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	140.0	18.9				mg/L	3	Standard
Dy	164	65719.4	1.8				mg/L	6	Standard
Ho-1	165	47698.3	1.5				mg/L	10	Standard
Er	166	45979.7	1.2				mg/L	13	Standard
I	127	5242.6	1.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		105.530	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.553	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138402
Report Date/Time: Wednesday, June 29, 2016 19:47:41
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	113.885
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.665
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606138402
Report Date/Time: Wednesday, June 29, 2016 19:47:41
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138404
Sample Date/Time: Wednesday, June 29, 2016 19:48:35
Number of Replicates: 3
Autosampler Position: 405
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80955.8	1.9				ug/L	74699	Standard
	Be	9	880.0	13.6	**0.9302**	0.126	13.6	ug/L	8	Standard
	Al	27	4591606.5	1.6	**28.8172**	0.235	0.8	ug/L	1493	Standard
	Sc	45	81773.4	1.2				ug/L	61029	Standard
	Ti	47	8450.7	1.8	**19.2499**	0.245	1.3	ug/L	35	Standard
	V	51	185955.1	1.6	**21.1827**	0.308	1.5	ug/L	1435	Standard
	Cr	52	431888.4	0.7	**49.5778**	0.407	0.8	ug/L	7635	Standard
	Cr	53	54036.2	1.9	**50.4433**	1.377	2.7	ug/L	363	Standard
	Mn	55	1979621.7	0.4	**246.7487**	3.142	1.3	ug/L	747	Standard
	Co	59	89211.3	2.3	**11.6929**	0.262	2.2	ug/L	234	Standard
	Ni	60	241828.7	1.1	**112.0223**	0.265	0.2	ug/L	215	Standard
	Cu	65	79842.9	1.5	**37.1424**	0.248	0.7	ug/L	104	Standard
	Zn	66	363879.1	1.7	**288.7674**	1.874	0.6	ug/L	363	Standard
>	**Ge**	72	570278.1	1.3				ug/L	495142	Standard
	As	75	11266.0	1.4	**8.7923**	0.037	0.4	ug/L	-50	Standard
	Se	82	562.8	1.7	**4.3149**	0.122	2.8	ug/L	14	Standard
	Se-1	77	468.7	10.3	**4.9506**	0.587	11.9	ug/L	42	Standard
>	**Ga**	71	6996.6	2.8				mg/L	42	Standard
	Rb	85	75201.3	3.4				ug/L	17	Standard
	Y	89	1637404.7	1.2				ug/L	417098	Standard
>	**Rh**	103	40.0	12.5				ug/L	13	Standard
	Mo	98	6065.7	2.4	**2.0076**	0.060	3.0	ug/L	11	Standard
	Ag	107	921.4	1.6	**0.1122**	0.003	2.7	ug/L	68	Standard
	Cd	111	1933.3	0.9	**0.8101**	0.015	1.9	mg/L	2	Standard
	Cd	114	4671.7	0.8	**0.8037**	0.006	0.7	ug/L	12	Standard
>	**In**	115	664089.4	1.0				ug/L	587441	Standard
	Sn	118	708.3	7.1	**0.0125**	0.007	53.2	ug/L	497	Standard
	Sb	123	171.2	9.0	**0.0282**	0.003	10.1	ug/L	78	Standard
	Ba	135	217409.8	1.0	**93.9488**	0.914	1.0	ug/L	24	Standard
	Ce	140	744689.2	0.9				ug/L	25	Standard
>	**Tb**	159	1077211.2	1.7				ug/L	939701	Standard
	Ho	165	62662.7	1.1				ug/L	10	Standard
	Tl	203	2865.3	1.3	**0.3204**	0.004	1.2	ug/L	4	Standard
	Tl	205	2535.2	5.2	**0.3286**	0.016	4.9	ug/L	2	Standard
	Pb	206	40598.7	0.9	**6.9743**	0.127	1.8	ug/L	240	Standard
	Pb	207	29975.3	2.2	**5.7854**	0.186	3.2	ug/L	198	Standard
	Pb	208	132785.8	0.6	**6.4904**	0.091	1.4	ug/L	883	Standard
	U	238	27584.4	0.2	**3.7106**	0.059	1.6	ug/L	3	Standard
>	**Bi**	209	565317.9	1.5				ug/L	523984	Standard

Sample ID: L1606138404
Report Date/Time: Wednesday, June 29, 2016 19:50:52
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**14.1527**	24.505	173.1	mg/L	0	Standard
Mg	24	243.3	11.7	**0.1691**	0.025	14.5	mg/L	37	Standard
K	39	461.7	5.5	**0.6478**	0.049	7.5	mg/L	38	Standard
Ca	43	205.0	4.9	**7.1444**	0.559	7.8	mg/L	10	Standard
Fe	54	23854.6	1.3	**14.9557**	0.059	0.4	mg/L	179	Standard
Fe	57	6316.3	1.9	**14.9472**	0.123	0.8	mg/L	335	Standard
> **Sc-1**	45	81773.4	1.2				mg/L	61029	Standard
Cl	35	26395.0	1.0				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	8435.7	8.5				ug/L	1547	Standard
P	31	20819.3	1.9				ug/L	15286	Standard
S	34	2495.2	1.6				ug/L	1572	Standard
Sr	88	186.7	12.1				ug/L	73	Standard
C	12	26.7	21.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	520.0	8.4				mg/L	3	Standard
Dy	164	85733.4	1.4				mg/L	6	Standard
Ho-1	165	62662.7	1.1				mg/L	10	Standard
Er	166	61347.3	2.7				mg/L	13	Standard
I	127	12962.2	2.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		108.376	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.175	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138404
Report Date/Time: Wednesday, June 29, 2016 19:50:52
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.048
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.888
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ni 60 Upper, S, EEE	Ni	60	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606138404
Report Date/Time: Wednesday, June 29, 2016 19:50:52
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 19:51:49
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	79632.1	2.5				ug/L	74699	Standard
	Be	9	53037.7	1.9	**56.5448**	1.909	3.4	ug/L	8	Standard
	Al	27	8288547.2	1.5	**52.9043**	0.899	1.7	ug/L	1493	Standard
	Sc	45	68251.7	0.4				ug/L	61029	Standard
	Ti	47	44871.7	1.2	**102.9405**	2.024	2.0	ug/L	35	Standard
	V	51	450952.2	1.7	**51.7787**	1.350	2.6	ug/L	1435	Standard
	Cr	52	441674.6	0.8	**50.8503**	0.824	1.6	ug/L	7635	Standard
	Cr	53	54163.4	3.4	**50.6767**	1.716	3.4	ug/L	363	Standard
	Mn	55	408648.7	3.2	**50.9466**	2.025	4.0	ug/L	747	Standard
	Co	59	404996.9	2.1	**53.3495**	1.595	3.0	ug/L	234	Standard
	Ni	60	109209.9	1.2	**50.6487**	1.081	2.1	ug/L	215	Standard
	Cu	65	108068.7	1.0	**50.4372**	0.990	2.0	ug/L	104	Standard
	Zn	66	63914.2	1.2	**50.4711**	0.538	1.1	ug/L	363	Standard
>	Ge	72	568913.3	1.0				ug/L	495142	Standard
	As	75	64596.5	1.7	**50.2573**	1.198	2.4	ug/L	-50	Standard
	Se	82	6434.6	2.8	**50.8949**	1.678	3.3	ug/L	14	Standard
	Se-1	77	4404.6	0.9	**51.8699**	0.916	1.8	ug/L	42	Standard
>	Ga	71	81.7	19.7				mg/L	42	Standard
	Rb	85	868.4	8.5				ug/L	17	Standard
	Y	89	484608.1	0.7				ug/L	417098	Standard
>	Rh	103	25.0	40.0				ug/L	13	Standard
	Mo	98	300990.3	2.1	**101.3882**	1.976	1.9	ug/L	11	Standard
	Ag	107	359716.1	0.6	**49.5455**	0.289	0.6	ug/L	68	Standard
	Cd	111	115140.5	0.5	**49.4382**	0.205	0.4	mg/L	2	Standard
	Cd	114	277323.3	2.2	**48.5511**	1.074	2.2	ug/L	12	Standard
>	In	115	654244.8	0.1				ug/L	587441	Standard
	Sn	118	326410.3	0.9	**50.0768**	0.438	0.9	ug/L	497	Standard
	Sb	123	257734.5	0.6	**49.5626**	0.244	0.5	ug/L	78	Standard
	Ba	135	107937.1	0.2	**47.3350**	0.075	0.2	ug/L	24	Standard
	Ce	140	156.7	33.8				ug/L	25	Standard
>	Tb	159	993573.1	1.0				ug/L	939701	Standard
	Ho	165	18.3	41.7				ug/L	10	Standard
	Tl	203	442343.3	0.6	**51.3703**	0.497	1.0	ug/L	4	Standard
	Tl	205	397679.6	3.2	**53.2177**	1.929	3.6	ug/L	2	Standard
	Pb	206	282572.7	0.8	**50.5712**	0.781	1.5	ug/L	240	Standard
	Pb	207	252130.0	1.0	**50.7898**	0.969	1.9	ug/L	198	Standard
	Pb	208	1034503.8	1.3	**52.7468**	1.180	2.2	ug/L	883	Standard
	U	238	372072.1	1.5	**51.7486**	1.271	2.5	ug/L	3	Standard
>	Bi	209	547364.2	0.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:54:05
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.1101**	29.627	173.2	mg/L	0	Standard
Mg	24	4437.3	1.9	**5.0568**	0.105	2.1	mg/L	37	Standard
K	39	2166.8	4.9	**4.1677**	0.193	4.6	mg/L	38	Standard
Ca	43	105.0	40.7	**3.4097**	2.426	71.2	mg/L	10	Standard
Fe	54	6964.2	0.4	**5.1415**	0.037	0.7	mg/L	179	Standard
Fe	57	2053.5	5.5	**5.1659**	0.318	6.1	mg/L	335	Standard
> **Sc-1**	45	68251.7	0.4				mg/L	61029	Standard
Cl	35	27215.8	2.7				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	1920.1	6.9				ug/L	1547	Standard
P	31	20039.9	1.0				ug/L	15286	Standard
S	34	2546.9	7.4				ug/L	1572	Standard
Sr	88	71.7	29.0				ug/L	73	Standard
C	12	70.0	24.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	11.7	100.8				mg/L	6	Standard
Ho-1	165	18.3	41.7				mg/L	10	Standard
Er	166	33.3	34.6				mg/L	13	Standard
I	127	1696.8	22.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	113.090		
Al	27	105.809		
Sc	45			
Ti	47	102.941		
V	51	103.557		
Cr	52	101.701		
Cr	53			
Mn	55	101.893		
Co	59	106.699		
Ni	60	101.297		
Cu	65	100.874		
Zn	66	100.942		
> Ge	72		114.899	
As	75	100.515		
Se	82	101.790		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:54:05
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.388		
Ag	107	99.091		
Cd	111	98.876		
Cd	114			
In	115		111.372	
Sn	118	100.154		
Sb	123	99.125		
Ba	135	94.670		
Ce	140			
Tb	159			
Ho	165			
Tl	203	102.741		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.494		
U	238	103.497		
Bi	209		104.462	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:54:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 19:55:00
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	78377.2	0.9				ug/L	74699	Standard
	Be	9	16.7	17.3	**0.0256**	0.003	11.6	ug/L	8	Standard
	Al	27	3955.5	8.5	**0.0123**	0.002	16.7	ug/L	1493	Standard
	Sc	45	70123.6	2.1				ug/L	61029	Standard
	Ti	47	44.3	19.2	**-0.0067**	0.022	331.3	ug/L	35	Standard
	V	51	1494.9	5.1	**-0.0231**	0.005	19.5	ug/L	1435	Standard
	Cr	52	7869.7	1.7	**-0.0976**	0.038	39.0	ug/L	7635	Standard
	Cr	53	438.3	4.0	**0.0839**	0.027	31.8	ug/L	363	Standard
	Mn	55	1323.4	11.3	**0.0175**	0.017	96.4	ug/L	747	Standard
	Co	59	292.0	15.1	**0.0012**	0.005	416.2	ug/L	234	Standard
	Ni	60	231.7	5.0	**-0.0160**	0.006	40.4	ug/L	215	Standard
	Cu	65	142.7	7.1	**-0.0369**	0.003	9.0	ug/L	104	Standard
	Zn	66	477.7	7.4	**-0.0736**	0.020	26.8	ug/L	363	Standard
>	Ge	72	568079.4	2.5				ug/L	495142	Standard
	As	75	-36.3	34.6	**0.0313**	0.010	31.8	ug/L	-50	Standard
	Se	82	17.1	26.2	**-0.0026**	0.033	1282.0	ug/L	14	Standard
	Se-1	77	54.3	9.3	**0.0268**	0.068	253.5	ug/L	42	Standard
>	Ga	71	40.0	12.5				mg/L	42	Standard
	Rb	85	38.3	19.9				ug/L	17	Standard
	Y	89	483139.1	0.9				ug/L	417098	Standard
>	Rh	103	11.7	99.0				ug/L	13	Standard
	Mo	98	291.7	36.0	**0.0908**	0.037	40.3	ug/L	11	Standard
	Ag	107	161.3	21.3	**0.0089**	0.005	57.1	ug/L	68	Standard
	Cd	111	21.8	28.2	**0.0012**	0.002	201.3	mg/L	2	Standard
	Cd	114	48.0	10.4	**0.0061**	0.001	15.7	ug/L	12	Standard
>	In	115	669206.2	1.8				ug/L	587441	Standard
	Sn	118	881.7	18.2	**0.0375**	0.022	58.1	ug/L	497	Standard
	Sb	123	447.3	19.7	**0.0797**	0.015	18.9	ug/L	78	Standard
	Ba	135	111.0	14.0	**0.0316**	0.007	21.6	ug/L	24	Standard
	Ce	140	185.0	42.5				ug/L	25	Standard
>	Tb	159	1027080.2	0.9				ug/L	939701	Standard
	Ho	165	28.3	10.2				ug/L	10	Standard
	Tl	203	102.0	27.3	**0.0095**	0.003	32.1	ug/L	4	Standard
	Tl	205	76.7	13.6	**0.0100**	0.001	13.1	ug/L	2	Standard
	Pb	206	326.3	10.0	**-0.0155**	0.005	35.3	ug/L	240	Standard
	Pb	207	292.3	4.6	**-0.0139**	0.002	17.4	ug/L	198	Standard
	Pb	208	1226.7	4.7	**-0.0145**	0.003	18.0	ug/L	883	Standard
	U	238	78.7	31.6	**0.0060**	0.003	54.4	ug/L	3	Standard
>	Bi	209	574845.6	0.5				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:57:17
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	34.4	**-0.0124**	0.019	155.4	mg/L	37	Standard
K	39	46.7	44.6	**-0.0238**	0.042	178.2	mg/L	38	Standard
Ca	43	26.7	57.3	**-1.0457**	0.868	83.0	mg/L	10	Standard
Fe	54	161.1	25.9	**-0.0190**	0.033	173.6	mg/L	179	Standard
Fe	57	251.7	19.3	**-0.3314**	0.129	38.8	mg/L	335	Standard
Sc-1	45	70123.6	2.1				mg/L	61029	Standard
Cl	35	27868.9	0.6				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	2040.1	1.8				ug/L	1547	Standard
P	31	20417.1	4.9				ug/L	15286	Standard
S	34	2425.2	3.5				ug/L	1572	Standard
Sr	88	45.0	40.1				ug/L	73	Standard
C	12	50.0	52.9				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	11.9	129.9				mg/L	6	Standard
Ho-1	165	28.3	10.2				mg/L	10	Standard
Er	166	30.0	57.7				mg/L	13	Standard
I	127	1460.1	42.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.731	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:57:17
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.919
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.707
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:57:17
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138406
Sample Date/Time: Wednesday, June 29, 2016 19:58:13
Number of Replicates: 3
Autosampler Position: 406
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77039.0	4.7				ug/L	74699	Standard
	Be	9	588.3	14.0	**0.6548**	0.074	11.3	ug/L	8	Standard
	Al	27	832772.7	0.7	**5.4883**	0.227	4.1	ug/L	1493	Standard
	Sc	45	75047.3	4.1				ug/L	61029	Standard
	Ti	47	6368.7	0.7	**14.6522**	0.311	2.1	ug/L	35	Standard
	V	51	122665.5	0.9	**14.0720**	0.287	2.0	ug/L	1435	Standard
	Cr	52	293324.5	0.5	**33.7422**	0.348	1.0	ug/L	7635	Standard
	Cr	53	35827.3	1.9	**33.7165**	0.271	0.8	ug/L	363	Standard
	Mn	55	1790039.2	2.5	**225.6481**	2.955	1.3	ug/L	747	Standard
	Co	59	59513.4	2.6	**7.8774**	0.097	1.2	ug/L	234	Standard
	Ni	60	162230.6	1.1	**75.9835**	0.734	1.0	ug/L	215	Standard
	Cu	65	58756.0	0.7	**27.6257**	0.286	1.0	ug/L	104	Standard
	Zn	66	213240.6	1.5	**171.0057**	1.221	0.7	ug/L	363	Standard
>	**Ge**	72	563753.3	1.5				ug/L	495142	Standard
	As	75	6354.9	1.3	**5.0439**	0.138	2.7	ug/L	-50	Standard
	Se	82	268.5	3.3	**2.0114**	0.102	5.1	ug/L	14	Standard
	Se-1	77	256.0	4.1	**2.4577**	0.168	6.8	ug/L	42	Standard
>	**Ga**	71	4010.5	2.8				mg/L	42	Standard
	Rb	85	39651.9	3.3				ug/L	17	Standard
	Y	89	1622063.0	2.1				ug/L	417098	Standard
>	**Rh**	103	33.3	8.7				ug/L	13	Standard
	Mo	98	4372.5	0.7	**1.4952**	0.034	2.3	ug/L	11	Standard
	Ag	107	608.3	2.2	**0.0725**	0.003	3.7	ug/L	68	Standard
	Cd	111	1369.4	3.4	**0.5916**	0.033	5.6	mg/L	2	Standard
	Cd	114	3220.4	7.1	**0.5731**	0.054	9.4	ug/L	12	Standard
>	**In**	115	642314.3	2.8				ug/L	587441	Standard
	Sn	118	866.7	3.3	**0.0411**	0.007	16.9	ug/L	497	Standard
	Sb	123	475.1	9.4	**0.0890**	0.010	11.7	ug/L	78	Standard
	Ba	135	210364.7	1.0	**94.0146**	1.722	1.8	ug/L	24	Standard
	Ce	140	620778.1	1.3				ug/L	25	Standard
>	**Tb**	159	1035282.0	2.1				ug/L	939701	Standard
	Ho	165	58499.4	0.8				ug/L	10	Standard
	Tl	203	1670.8	3.4	**0.1930**	0.007	3.5	ug/L	4	Standard
	Tl	205	1583.4	5.7	**0.2129**	0.013	6.3	ug/L	2	Standard
	Pb	206	23886.1	1.1	**4.2252**	0.039	0.9	ug/L	240	Standard
	Pb	207	17508.1	0.9	**3.4746**	0.041	1.2	ug/L	198	Standard
	Pb	208	79776.0	1.1	**4.0140**	0.064	1.6	ug/L	883	Standard
	U	238	22400.5	0.6	**3.1225**	0.026	0.8	ug/L	3	Standard
>	**Bi**	209	545342.1	0.8				ug/L	523984	Standard

Sample ID: L1606138406
Report Date/Time: Wednesday, June 29, 2016 20:00:30
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	190.0	5.3	**0.1344**	0.009	6.3	mg/L	37	Standard
K	39	241.7	12.5	**0.3203**	0.051	16.0	mg/L	38	Standard
Ca	43	168.3	3.4	**6.1267**	0.214	3.5	mg/L	10	Standard
Fe	54	12517.1	3.0	**8.4992**	0.364	4.3	mg/L	179	Standard
Fe	57	3428.7	4.4	**8.4118**	0.531	6.3	mg/L	335	Standard
> **Sc-1**	45	75047.3	4.1				mg/L	61029	Standard
Cl	35	26434.4	2.8				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2270.2	3.9				ug/L	1547	Standard
P	31	19440.8	0.5				ug/L	15286	Standard
S	34	2256.8	6.7				ug/L	1572	Standard
Sr	88	170.0	33.4				ug/L	73	Standard
C	12	76.7	54.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	300.0	18.6				mg/L	3	Standard
Dy	164	81135.2	0.1				mg/L	6	Standard
Ho-1	165	58499.4	0.8				mg/L	10	Standard
Er	166	56983.6	0.5				mg/L	13	Standard
I	127	5462.7	4.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.133	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.857	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138406
Report Date/Time: Wednesday, June 29, 2016 20:00:30
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	109.341
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.076
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138408
Sample Date/Time: Wednesday, June 29, 2016 20:01:24
Number of Replicates: 3
Autosampler Position: 407
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79667.3	1.9				ug/L	74699	Standard
	Be	9	540.0	10.4	**0.5836**	0.070	11.9	ug/L	8	Standard
	Al	27	892686.6	1.0	**5.6842**	0.164	2.9	ug/L	1493	Standard
	Sc	45	78111.0	3.7				ug/L	61029	Standard
	Ti	47	6171.3	0.9	**13.8487**	0.141	1.0	ug/L	35	Standard
	V	51	131186.8	0.8	**14.6946**	0.149	1.0	ug/L	1435	Standard
	Cr	52	291322.4	2.4	**32.6724**	0.743	2.3	ug/L	7635	Standard
	Cr	53	35120.7	4.3	**32.2467**	1.471	4.6	ug/L	363	Standard
	Mn	55	2233691.8	2.2	**274.8758**	6.127	2.2	ug/L	747	Standard
	Co	59	65245.2	1.7	**8.4325**	0.147	1.7	ug/L	234	Standard
	Ni	60	190260.1	2.2	**86.9887**	2.023	2.3	ug/L	215	Standard
	Cu	65	66483.6	1.3	**30.5171**	0.399	1.3	ug/L	104	Standard
	Zn	66	187493.0	1.2	**146.6893**	2.307	1.6	ug/L	363	Standard
>	**Ge**	72	577612.7	0.6				ug/L	495142	Standard
	As	75	6445.8	1.9	**4.9929**	0.116	2.3	ug/L	-50	Standard
	Se	82	222.3	2.5	**1.5982**	0.037	2.3	ug/L	14	Standard
	Se-1	77	243.0	1.1	**2.2298**	0.039	1.8	ug/L	42	Standard
>	**Ga**	71	4360.6	1.0				mg/L	42	Standard
	Rb	85	41170.9	1.0				ug/L	17	Standard
	Y	89	1636669.5	0.9				ug/L	417098	Standard
>	**Rh**	103	33.3	22.9				ug/L	13	Standard
	Mo	98	4750.5	1.5	**1.5922**	0.009	0.6	ug/L	11	Standard
	Ag	107	651.7	5.1	**0.0767**	0.003	4.5	ug/L	68	Standard
	Cd	111	1476.5	1.8	**0.6252**	0.009	1.4	mg/L	2	Standard
	Cd	114	3537.9	6.1	**0.6161**	0.029	4.7	ug/L	12	Standard
>	**In**	115	655238.7	1.5				ug/L	587441	Standard
	Sn	118	965.0	3.4	**0.0534**	0.004	7.7	ug/L	497	Standard
	Sb	123	268.7	11.3	**0.0474**	0.005	10.8	ug/L	78	Standard
	Ba	135	196413.8	1.6	**86.0176**	0.137	0.2	ug/L	24	Standard
	Ce	140	596487.3	1.9				ug/L	25	Standard
>	**Tb**	159	1073225.5	1.0				ug/L	939701	Standard
	Ho	165	61178.3	1.7				ug/L	10	Standard
	Tl	203	2456.5	1.8	**0.2718**	0.001	0.3	ug/L	4	Standard
	Tl	205	2150.2	4.8	**0.2763**	0.017	6.1	ug/L	2	Standard
	Pb	206	24200.3	1.9	**4.0880**	0.010	0.3	ug/L	240	Standard
	Pb	207	17656.3	0.7	**3.3464**	0.078	2.3	ug/L	198	Standard
	Pb	208	79187.0	3.1	**3.8025**	0.046	1.2	ug/L	883	Standard
	U	238	22940.4	1.4	**3.0557**	0.038	1.2	ug/L	3	Standard
>	**Bi**	209	570736.3	2.0				ug/L	523984	Standard

Sample ID: L1606138408
Report Date/Time: Wednesday, June 29, 2016 20:03:41
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**28.7877**	49.853	173.2	mg/L	0	Standard
Mg	24	161.7	15.3	**0.0975**	0.019	19.1	mg/L	37	Standard
K	39	273.3	12.8	**0.3583**	0.064	17.7	mg/L	38	Standard
Ca	43	191.7	9.9	**6.9329**	0.825	11.9	mg/L	10	Standard
Fe	54	12889.7	4.6	**8.4015**	0.298	3.5	mg/L	179	Standard
Fe	57	3420.4	5.7	**8.0292**	0.829	10.3	mg/L	335	Standard
> **Sc-1**	45	78111.0	3.7				mg/L	61029	Standard
Cl	35	25108.8	1.1				ug/L	20195	Standard
Kr	83	1.0	173.2				ug/L	4	Standard
Br	81	1990.1	4.1				ug/L	1547	Standard
P	31	18960.2	2.1				ug/L	15286	Standard
S	34	2216.8	7.1				ug/L	1572	Standard
Sr	88	153.3	28.5				ug/L	73	Standard
C	12	110.0	9.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	313.3	12.9				mg/L	3	Standard
Dy	164	83986.8	2.5				mg/L	6	Standard
Ho-1	165	61178.3	1.7				mg/L	10	Standard
Er	166	57776.6	0.9				mg/L	13	Standard
I	127	4875.8	1.4				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		106.651	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		116.656	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606138408
Report Date/Time: Wednesday, June 29, 2016 20:03:41
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	111.541
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.922
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138410
Sample Date/Time: Wednesday, June 29, 2016 20:04:36
Number of Replicates: 3
Autosampler Position: 408
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82222.6	2.1				ug/L	74699	Standard
	Be	9	466.7	0.6	**0.4893**	0.010	2.0	ug/L	8	Standard
	Al	27	798979.6	1.9	**4.9264**	0.080	1.6	ug/L	1493	Standard
	Sc	45	79640.5	1.7				ug/L	61029	Standard
	Ti	47	5792.4	2.5	**13.1770**	0.593	4.5	ug/L	35	Standard
	V	51	95139.7	1.4	**10.7518**	0.246	2.3	ug/L	1435	Standard
	Cr	52	279542.5	2.0	**31.7533**	0.664	2.1	ug/L	7635	Standard
	Cr	53	34038.2	1.3	**31.6794**	0.880	2.8	ug/L	363	Standard
	Mn	55	1053790.0	0.3	**131.3975**	3.080	2.3	ug/L	747	Standard
	Co	59	35027.1	0.6	**4.5723**	0.080	1.8	ug/L	234	Standard
	Ni	60	156240.8	1.2	**72.3985**	1.638	2.3	ug/L	215	Standard
	Cu	65	48550.0	0.3	**22.5657**	0.405	1.8	ug/L	104	Standard
	Zn	66	257712.2	0.3	**204.5897**	4.503	2.2	ug/L	363	Standard
>	**Ge**	72	569889.2	2.0				ug/L	495142	Standard
	As	75	6132.6	2.6	**4.8186**	0.190	4.0	ug/L	-50	Standard
	Se	82	361.1	9.2	**2.7244**	0.314	11.5	ug/L	14	Standard
	Se-1	77	331.7	7.4	**3.3212**	0.233	7.0	ug/L	42	Standard
>	**Ga**	71	4152.2	2.1				mg/L	42	Standard
	Rb	85	52623.0	3.5				ug/L	17	Standard
	Y	89	1425089.1	0.3				ug/L	417098	Standard
>	**Rh**	103	33.3	8.7				ug/L	13	Standard
	Mo	98	3927.9	0.3	**1.2648**	0.018	1.4	ug/L	11	Standard
	Ag	107	467.0	2.9	**0.0489**	0.003	5.2	ug/L	68	Standard
	Cd	111	1169.0	3.4	**0.4740**	0.011	2.3	mg/L	2	Standard
	Cd	114	2891.7	1.0	**0.4840**	0.007	1.5	ug/L	12	Standard
>	**In**	115	681463.1	1.2				ug/L	587441	Standard
	Sn	118	755.0	5.0	**0.0168**	0.007	41.1	ug/L	497	Standard
	Sb	123	219.3	13.0	**0.0363**	0.005	13.2	ug/L	78	Standard
	Ba	135	73803.3	1.7	**31.0670**	0.270	0.9	ug/L	24	Standard
	Ce	140	554709.1	1.1				ug/L	25	Standard
>	**Tb**	159	1094015.2	1.4				ug/L	939701	Standard
	Ho	165	50373.6	1.6				ug/L	10	Standard
	Tl	203	1120.7	1.3	**0.1214**	0.002	2.0	ug/L	4	Standard
	Tl	205	941.7	7.2	**0.1194**	0.008	6.7	ug/L	2	Standard
	Pb	206	24028.0	1.2	**4.0020**	0.045	1.1	ug/L	240	Standard
	Pb	207	17504.4	0.3	**3.2701**	0.054	1.6	ug/L	198	Standard
	Pb	208	78967.6	0.8	**3.7402**	0.091	2.4	ug/L	883	Standard
	U	238	25248.7	0.7	**3.3178**	0.077	2.3	ug/L	3	Standard
>	**Bi**	209	578682.9	1.6				ug/L	523984	Standard

Sample ID: L1606138410
Report Date/Time: Wednesday, June 29, 2016 20:06:53
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	188.3	33.0	**0.1211**	0.061	50.1	mg/L	37	Standard
K	39	248.3	3.1	**0.3067**	0.020	6.4	mg/L	38	Standard
Ca	43	173.3	7.3	**5.8747**	0.743	12.6	mg/L	10	Standard
Fe	54	14056.9	1.9	**8.9942**	0.034	0.4	mg/L	179	Standard
Fe	57	3822.2	6.9	**8.8892**	0.835	9.4	mg/L	335	Standard
> **Sc-1**	45	79640.5	1.7				mg/L	61029	Standard
Cl	35	25424.6	1.4				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	2053.5	6.1				ug/L	1547	Standard
P	31	20016.5	1.3				ug/L	15286	Standard
S	34	2301.8	5.1				ug/L	1572	Standard
Sr	88	141.7	32.0				ug/L	73	Standard
C	12	56.7	36.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	133.3	11.5				mg/L	3	Standard
Dy	164	72545.2	2.7				mg/L	6	Standard
Ho-1	165	50373.6	1.6				mg/L	10	Standard
Er	166	47160.0	0.9				mg/L	13	Standard
I	127	6799.9	2.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.072	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.096	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138410
Report Date/Time: Wednesday, June 29, 2016 20:06:53
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	116.005
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	110.439
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606138410
Report Date/Time: Wednesday, June 29, 2016 20:06:53
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138412
Sample Date/Time: Wednesday, June 29, 2016 20:07:47
Number of Replicates: 3
Autosampler Position: 409
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81391.5	3.4				ug/L	74699	Standard
	Be	9	565.0	6.9	**0.5962**	0.020	3.4	ug/L	8	Standard
	Al	27	802320.4	2.6	**4.9984**	0.082	1.6	ug/L	1493	Standard
	Sc	45	77871.4	3.1				ug/L	61029	Standard
	Ti	47	8431.3	2.1	**19.3525**	0.252	1.3	ug/L	35	Standard
	V	51	136034.9	0.6	**15.5641**	0.178	1.1	ug/L	1435	Standard
	Cr	52	292890.6	0.7	**33.5584**	0.660	2.0	ug/L	7635	Standard
	Cr	53	35682.0	1.9	**33.4491**	0.599	1.8	ug/L	363	Standard
	Mn	55	2260782.0	0.5	**283.9966**	5.862	2.1	ug/L	747	Standard
	Co	59	82153.8	0.1	**10.8487**	0.184	1.7	ug/L	234	Standard
	Ni	60	172794.1	1.0	**80.6313**	1.536	1.9	ug/L	215	Standard
	Cu	65	63565.7	0.4	**29.7801**	0.432	1.5	ug/L	104	Standard
	Zn	66	251019.4	0.8	**200.6145**	2.591	1.3	ug/L	363	Standard
>	**Ge**	72	565970.9	1.7				ug/L	495142	Standard
	As	75	7636.9	2.3	**6.0250**	0.142	2.4	ug/L	-50	Standard
	Se	82	426.3	0.8	**3.2611**	0.084	2.6	ug/L	14	Standard
	Se-1	77	381.7	1.3	**3.9506**	0.115	2.9	ug/L	42	Standard
>	**Ga**	71	4992.5	2.6				mg/L	42	Standard
	Rb	85	46121.8	0.5				ug/L	17	Standard
	Y	89	1754842.6	1.7				ug/L	417098	Standard
>	**Rh**	103	43.3	35.3				ug/L	13	Standard
	Mo	98	5382.6	1.1	**1.7465**	0.044	2.5	ug/L	11	Standard
	Ag	107	637.3	4.9	**0.0719**	0.004	5.2	ug/L	68	Standard
	Cd	111	1430.5	6.2	**0.5853**	0.028	4.9	mg/L	2	Standard
	Cd	114	3407.8	7.3	**0.5746**	0.047	8.2	ug/L	12	Standard
>	**In**	115	677210.7	1.5				ug/L	587441	Standard
	Sn	118	780.0	4.5	**0.0212**	0.006	30.1	ug/L	497	Standard
	Sb	123	243.4	9.9	**0.0411**	0.005	12.6	ug/L	78	Standard
	Ba	135	199262.8	0.6	**84.4450**	1.189	1.4	ug/L	24	Standard
	Ce	140	678125.1	0.7				ug/L	25	Standard
>	**Tb**	159	1071624.9	1.1				ug/L	939701	Standard
	Ho	165	67051.4	0.5				ug/L	10	Standard
	Tl	203	1859.8	4.4	**0.2096**	0.007	3.5	ug/L	4	Standard
	Tl	205	1828.4	5.2	**0.2396**	0.011	4.7	ug/L	2	Standard
	Pb	206	33182.6	0.6	**5.7494**	0.057	1.0	ug/L	240	Standard
	Pb	207	24419.7	0.3	**4.7511**	0.035	0.7	ug/L	198	Standard
	Pb	208	111320.6	0.3	**5.4889**	0.050	0.9	ug/L	883	Standard
	U	238	25349.9	1.3	**3.4462**	0.014	0.4	ug/L	3	Standard
>	**Bi**	209	559224.1	1.0				ug/L	523984	Standard

Sample ID: L1606138412
Report Date/Time: Wednesday, June 29, 2016 20:10:04
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**30.2574**	26.235	86.7	mg/L	0	Standard
Mg	24	168.3	9.1	**0.1052**	0.014	12.9	mg/L	37	Standard
K	39	280.0	6.4	**0.3709**	0.027	7.3	mg/L	38	Standard
Ca	43	176.7	27.2	**6.2736**	2.632	42.0	mg/L	10	Standard
Fe	54	15625.1	2.7	**10.2448**	0.046	0.4	mg/L	179	Standard
Fe	57	4357.3	3.7	**10.5343**	0.341	3.2	mg/L	335	Standard
> **Sc-1**	45	77871.4	3.1				mg/L	61029	Standard
Cl	35	26050.4	2.2				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2310.2	5.7				ug/L	1547	Standard
P	31	19966.5	2.1				ug/L	15286	Standard
S	34	2301.8	1.2				ug/L	1572	Standard
Sr	88	181.7	8.4				ug/L	73	Standard
C	12	60.0	66.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	466.7	15.8				mg/L	3	Standard
Dy	164	94635.2	0.7				mg/L	6	Standard
Ho-1	165	67051.4	0.5				mg/L	10	Standard
Er	166	62943.8	1.5				mg/L	13	Standard
I	127	6689.8	6.1				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		108.960	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		114.305	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606138412
Report Date/Time: Wednesday, June 29, 2016 20:10:04
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	115.281
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	106.725
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 20:11:01
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	78720.7	2.9				ug/L	74699	Standard
	Be	9	53925.8	1.9	**58.1477**	1.206	2.1	ug/L	8	Standard
	Al	27	8443805.5	1.2	**54.5410**	1.847	3.4	ug/L	1493	Standard
	Sc	45	68074.4	3.5				ug/L	61029	Standard
	Ti	47	45146.2	1.1	**105.2966**	1.372	1.3	ug/L	35	Standard
	V	51	452407.6	0.9	**52.8112**	0.533	1.0	ug/L	1435	Standard
	Cr	52	439865.6	0.2	**51.4989**	0.494	1.0	ug/L	7635	Standard
	Cr	53	55131.8	1.5	**52.4610**	1.191	2.3	ug/L	363	Standard
	Mn	55	411766.6	1.7	**52.1886**	1.083	2.1	ug/L	747	Standard
	Co	59	411137.6	1.8	**55.0576**	1.090	2.0	ug/L	234	Standard
	Ni	60	108563.6	1.0	**51.1869**	0.633	1.2	ug/L	215	Standard
	Cu	65	108514.2	0.8	**51.4879**	0.273	0.5	ug/L	104	Standard
	Zn	66	63813.1	1.5	**51.2390**	0.656	1.3	ug/L	363	Standard
>	Ge	72	559563.9	0.9				ug/L	495142	Standard
	As	75	65025.3	0.7	**51.4299**	0.351	0.7	ug/L	-50	Standard
	Se	82	6607.2	0.8	**53.1327**	0.267	0.5	ug/L	14	Standard
	Se-1	77	4509.3	0.5	**54.0117**	0.383	0.7	ug/L	42	Standard
>	Ga	71	68.3	11.2				mg/L	42	Standard
	Rb	85	971.7	11.9				ug/L	17	Standard
	Y	89	484276.8	2.3				ug/L	417098	Standard
>	Rh	103	33.3	52.7				ug/L	13	Standard
	Mo	98	305352.5	1.7	**102.8803**	2.601	2.5	ug/L	11	Standard
	Ag	107	366186.0	1.2	**50.4411**	0.533	1.1	ug/L	68	Standard
	Cd	111	118185.8	1.5	**50.7502**	0.681	1.3	mg/L	2	Standard
	Cd	114	288421.2	1.0	**50.5001**	0.711	1.4	ug/L	12	Standard
>	In	115	654225.0	1.4				ug/L	587441	Standard
	Sn	118	325668.1	1.2	**49.9654**	0.174	0.3	ug/L	497	Standard
	Sb	123	260726.4	2.1	**50.1401**	0.797	1.6	ug/L	78	Standard
	Ba	135	109583.9	1.3	**48.0616**	0.591	1.2	ug/L	24	Standard
	Ce	140	191.7	20.9				ug/L	25	Standard
>	Tb	159	983539.2	1.4				ug/L	939701	Standard
	Ho	165	21.7	13.3				ug/L	10	Standard
	Tl	203	449254.6	0.7	**52.6820**	0.493	0.9	ug/L	4	Standard
	Tl	205	410287.0	1.8	**55.4353**	1.032	1.9	ug/L	2	Standard
	Pb	206	287176.7	1.5	**51.8952**	0.724	1.4	ug/L	240	Standard
	Pb	207	254226.6	1.1	**51.7082**	0.492	1.0	ug/L	198	Standard
	Pb	208	1062593.6	0.7	**54.7051**	0.438	0.8	ug/L	883	Standard
	U	238	383746.1	0.8	**53.8863**	0.314	0.6	ug/L	3	Standard
>	Bi	209	542059.3	0.3				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:13:18
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.9421**	31.068	173.2	mg/L	0	Standard
Mg	24	4572.4	3.8	**5.2291**	0.219	4.2	mg/L	37	Standard
K	39	2235.2	5.1	**4.3138**	0.121	2.8	mg/L	38	Standard
Ca	43	108.3	35.8	**3.6601**	2.380	65.0	mg/L	10	Standard
Fe	54	7126.2	2.9	**5.2851**	0.313	5.9	mg/L	179	Standard
Fe	57	2081.8	4.0	**5.2741**	0.321	6.1	mg/L	335	Standard
Sc-1	45	68074.4	3.5				mg/L	61029	Standard
Cl	35	27726.7	0.9				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	1910.1	6.0				ug/L	1547	Standard
P	31	20502.2	2.0				ug/L	15286	Standard
S	34	2426.9	4.0				ug/L	1572	Standard
Sr	88	50.0	26.5				ug/L	73	Standard
C	12	73.3	20.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	15.9	37.3				mg/L	6	Standard
Ho-1	165	21.7	13.3				mg/L	10	Standard
Er	166	16.7	34.6				mg/L	13	Standard
I	127	1525.1	31.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	116.295		
Al	27	109.082		
Sc	45			
Ti	47	105.297		
V	51	105.622		
Cr	52	102.998		
Cr	53			
Mn	55	104.377		
Co	59	110.115		
Ni	60	102.374		
Cu	65	102.976		
Zn	66	102.478		
Ge	72		113.011	
As	75	102.860		
Se	82	106.265		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:13:18
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.880		
Ag	107	100.882		
Cd	111	101.500		
Cd	114			
In	115		111.369	
Sn	118	99.931		
Sb	123	100.280		
Ba	135	96.123		
Ce	140			
Tb	159			
Ho	165			
Tl	203	105.364		
Tl	205			
Pb	206			
Pb	207			
Pb	208	109.410		
U	238	107.773		
Bi	209		103.450	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:13:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 20:14:12
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79194.8	0.3				ug/L	74699	Standard
	Be	9	31.7	9.1	**0.0415**	0.003	7.4	ug/L	8	Standard
	Al	27	5331.0	39.7	**0.0208**	0.013	64.7	ug/L	1493	Standard
	Sc	45	70443.5	2.7				ug/L	61029	Standard
	Ti	47	45.0	15.6	**-0.0077**	0.016	206.1	ug/L	35	Standard
	V	51	1613.4	5.4	**-0.0133**	0.013	95.3	ug/L	1435	Standard
	Cr	52	7981.4	3.6	**-0.1047**	0.047	45.2	ug/L	7635	Standard
	Cr	53	418.3	25.3	**0.0570**	0.103	180.4	ug/L	363	Standard
	Mn	55	1741.1	30.3	**0.0658**	0.068	103.0	ug/L	747	Standard
	Co	59	378.7	17.0	**0.0117**	0.009	74.2	ug/L	234	Standard
	Ni	60	306.7	29.3	**0.0162**	0.043	265.9	ug/L	215	Standard
	Cu	65	193.0	22.7	**-0.0150**	0.021	143.0	ug/L	104	Standard
	Zn	66	576.0	19.1	**-0.0039**	0.093	2387.7	ug/L	363	Standard
>	**Ge**	72	580508.2	1.5				ug/L	495142	Standard
	As	75	-17.3	276.0	**0.0468**	0.036	77.6	ug/L	-50	Standard
	Se	82	19.7	47.6	**0.0162**	0.076	468.0	ug/L	14	Standard
	Se-1	77	62.3	8.8	**0.1062**	0.073	68.4	ug/L	42	Standard
>	**Ga**	71	43.3	24.0				mg/L	42	Standard
	Rb	85	53.3	27.1				ug/L	17	Standard
	Y	89	466004.6	0.7				ug/L	417098	Standard
>	**Rh**	103	18.3	41.7				ug/L	13	Standard
	Mo	98	326.5	17.1	**0.1045**	0.019	18.6	ug/L	11	Standard
	Ag	107	195.7	20.5	**0.0141**	0.005	38.2	ug/L	68	Standard
	Cd	111	49.8	49.3	**0.0134**	0.010	76.8	mg/L	2	Standard
	Cd	114	97.6	30.0	**0.0150**	0.005	33.6	ug/L	12	Standard
>	**In**	115	653349.0	0.9				ug/L	587441	Standard
	Sn	118	958.4	16.5	**0.0527**	0.023	43.7	ug/L	497	Standard
	Sb	123	468.2	24.3	**0.0858**	0.021	24.7	ug/L	78	Standard
	Ba	135	169.0	53.9	**0.0580**	0.039	67.5	ug/L	24	Standard
	Ce	140	383.3	72.7				ug/L	25	Standard
>	**Tb**	159	1008003.3	2.5				ug/L	939701	Standard
	Ho	165	33.3	56.8				ug/L	10	Standard
	Tl	203	138.0	27.1	**0.0136**	0.004	29.9	ug/L	4	Standard
	Tl	205	128.3	47.6	**0.0167**	0.008	47.6	ug/L	2	Standard
	Pb	206	364.7	12.5	**-0.0084**	0.008	93.5	ug/L	240	Standard
	Pb	207	321.7	11.0	**-0.0076**	0.006	84.7	ug/L	198	Standard
	Pb	208	1298.7	10.5	**-0.0103**	0.007	67.0	ug/L	883	Standard
	U	238	121.3	35.3	**0.0118**	0.006	48.9	ug/L	3	Standard
>	**Bi**	209	569222.6	0.7				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:16:29
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	31.7	50.8	**-0.0295**	0.019	63.3	mg/L	37	Standard
K	39	58.3	13.1	**-0.0021**	0.017	817.3	mg/L	38	Standard
Ca	43	26.7	28.6	**-1.0586**	0.442	41.7	mg/L	10	Standard
Fe	54	154.2	11.8	**-0.0252**	0.011	43.9	mg/L	179	Standard
Fe	57	258.3	4.9	**-0.3125**	0.056	18.0	mg/L	335	Standard
Sc-1	45	70443.5	2.7				mg/L	61029	Standard
Cl	35	27476.9	1.3				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	1946.8	3.9				ug/L	1547	Standard
P	31	20592.3	4.9				ug/L	15286	Standard
S	34	2348.5	2.8				ug/L	1572	Standard
Sr	88	75.0	46.7				ug/L	73	Standard
C	12	43.3	13.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	74.1	39.2				mg/L	6	Standard
Ho-1	165	33.3	56.8				mg/L	10	Standard
Er	166	53.3	71.0				mg/L	13	Standard
I	127	1293.4	30.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		117.241	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:16:29
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.219
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.633
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:16:29
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBW 43 WG573291-03
Sample Date/Time: Wednesday, June 29, 2016 20:17:26
Number of Replicates: 3
Autosampler Position: 410
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78266.6	1.0				ug/L	74699	Standard
	Be	9	21.7	58.1	**0.0312**	0.014	44.4	ug/L	8	Standard
	Al	27	32901.2	14.4	**0.2003**	0.031	15.3	ug/L	1493	Standard
	Sc	45	68990.1	1.1				ug/L	61029	Standard
	Ti	47	59.7	25.6	**0.0271**	0.036	131.8	ug/L	35	Standard
	V	51	1740.6	13.9	**0.0036**	0.029	816.2	ug/L	1435	Standard
	Cr	52	10743.1	2.3	**0.2282**	0.033	14.5	ug/L	7635	Standard
	Cr	53	856.7	10.0	**0.4710**	0.086	18.2	ug/L	363	Standard
	Mn	55	4965.9	44.9	**0.4694**	0.280	59.7	ug/L	747	Standard
	Co	59	438.7	27.8	**0.0202**	0.016	81.0	ug/L	234	Standard
	Ni	60	464.7	17.9	**0.0906**	0.040	44.1	ug/L	215	Standard
	Cu	65	288.0	10.4	**0.0300**	0.015	49.7	ug/L	104	Standard
	Zn	66	2615.6	1.4	**1.6135**	0.041	2.5	ug/L	363	Standard
>	**Ge**	72	573366.1	0.8				ug/L	495142	Standard
	As	75	-37.9	81.3	**0.0303**	0.024	78.8	ug/L	-50	Standard
	Se	82	11.3	32.8	**-0.0488**	0.029	60.5	ug/L	14	Standard
	Se-1	77	61.0	9.1	**0.0990**	0.067	67.3	ug/L	42	Standard
>	**Ga**	71	31.7	9.1				mg/L	42	Standard
	Rb	85	100.0	26.5				ug/L	17	Standard
	Y	89	479627.1	2.2				ug/L	417098	Standard
>	**Rh**	103	1.7	173.2				ug/L	13	Standard
	Mo	98	155.9	47.8	**0.0477**	0.025	52.9	ug/L	11	Standard
	Ag	107	191.7	76.1	**0.0140**	0.020	145.7	ug/L	68	Standard
	Cd	111	21.5	58.8	**0.0014**	0.005	377.2	mg/L	2	Standard
	Cd	114	53.4	80.4	**0.0074**	0.008	102.3	ug/L	12	Standard
>	**In**	115	641242.2	1.6				ug/L	587441	Standard
	Sn	118	930.0	4.7	**0.0512**	0.008	15.1	ug/L	497	Standard
	Sb	123	286.5	19.6	**0.0520**	0.011	21.3	ug/L	78	Standard
	Ba	135	402.7	93.5	**0.1636**	0.167	102.4	ug/L	24	Standard
	Ce	140	951.7	120.0				ug/L	25	Standard
>	**Tb**	159	992014.1	1.0				ug/L	939701	Standard
	Ho	165	68.3	105.1				ug/L	10	Standard
	Tl	203	138.0	108.8	**0.0140**	0.017	122.9	ug/L	4	Standard
	Tl	205	136.7	67.7	**0.0182**	0.012	67.1	ug/L	2	Standard
	Pb	206	481.7	43.4	**0.0137**	0.037	272.5	ug/L	240	Standard
	Pb	207	395.0	43.8	**0.0084**	0.035	416.2	ug/L	198	Standard
	Pb	208	1692.0	26.2	**0.0108**	0.023	209.9	ug/L	883	Standard
	U	238	120.3	76.5	**0.0121**	0.013	104.7	ug/L	3	Standard
>	**Bi**	209	557465.1	0.7				ug/L	523984	Standard

Sample ID: PBW 43 WG573291-03
Report Date/Time: Wednesday, June 29, 2016 20:19:42
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	43.3	**-0.0120**	0.023	188.7	mg/L	37	Standard
K	39	41.7	73.3	**-0.0331**	0.059	178.0	mg/L	38	Standard
Ca	43	16.7	45.8	**-1.5897**	0.437	27.5	mg/L	10	Standard
Fe	54	190.8	15.9	**0.0047**	0.022	466.7	mg/L	179	Standard
Fe	57	281.7	18.5	**-0.2277**	0.154	67.5	mg/L	335	Standard
Sc-1	45	68990.1	1.1				mg/L	61029	Standard
Cl	35	27241.8	2.0				ug/L	20195	Standard
Kr	83	3.3	34.6				ug/L	4	Standard
Br	81	1960.1	2.7				ug/L	1547	Standard
P	31	19042.0	2.0				ug/L	15286	Standard
S	34	2271.8	2.5				ug/L	1572	Standard
Sr	88	70.0	14.3				ug/L	73	Standard
C	12	93.3	12.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	57.4	72.5				mg/L	6	Standard
Ho-1	165	68.3	105.1				mg/L	10	Standard
Er	166	53.3	75.8				mg/L	13	Standard
I	127	1241.7	37.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		104.776	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.798	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: PBW 43 WG573291-03
Report Date/Time: Wednesday, June 29, 2016 20:19:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.159
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.390
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 43 WG573291-03
Report Date/Time: Wednesday, June 29, 2016 20:19:42
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSW 43 WG573291-04
Sample Date/Time: Wednesday, June 29, 2016 20:20:37
Number of Replicates: 3
Autosampler Position: 411
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79286.9	1.2				ug/L	74699	Standard
	Be	9	55518.2	1.7	**59.4178**	0.565	1.0	ug/L	8	Standard
	Al	27	18234.3	3.1	**0.1035**	0.005	4.4	ug/L	1493	Standard
	Sc	45	71925.5	3.7				ug/L	61029	Standard
	Ti	47	66.7	12.9	**0.0416**	0.023	54.2	ug/L	35	Standard
	V	51	464385.2	1.4	**52.3659**	1.151	2.2	ug/L	1435	Standard
	Cr	52	468709.1	0.6	**53.0500**	1.578	3.0	ug/L	7635	Standard
	Cr	53	57400.2	0.6	**52.7689**	1.471	2.8	ug/L	363	Standard
	Mn	55	425738.4	0.7	**52.1249**	1.035	2.0	ug/L	747	Standard
	Co	59	411335.9	1.2	**53.2163**	1.483	2.8	ug/L	234	Standard
	Ni	60	114540.7	1.0	**52.1842**	1.705	3.3	ug/L	215	Standard
	Cu	65	115424.0	0.4	**52.9151**	1.312	2.5	ug/L	104	Standard
	Zn	66	68318.0	0.2	**53.0163**	1.377	2.6	ug/L	363	Standard
>	**Ge**	72	579422.8	2.7				ug/L	495142	Standard
	As	75	63888.1	1.5	**48.8227**	1.412	2.9	ug/L	-50	Standard
	Se	82	6548.4	1.5	**50.8872**	2.152	4.2	ug/L	14	Standard
	Se-1	77	4570.0	2.1	**52.8570**	0.847	1.6	ug/L	42	Standard
>	**Ga**	71	68.3	27.7				mg/L	42	Standard
	Rb	85	96.7	7.9				ug/L	17	Standard
	Y	89	471340.1	1.5				ug/L	417098	Standard
>	**Rh**	103	33.3	52.7				ug/L	13	Standard
	Mo	98	1844.9	1.6	**0.6204**	0.021	3.4	ug/L	11	Standard
	Ag	107	386699.9	0.4	**53.6468**	1.108	2.1	ug/L	68	Standard
	Cd	111	120754.4	0.7	**52.2256**	1.320	2.5	mg/L	2	Standard
	Cd	114	290744.9	1.2	**51.2713**	1.454	2.8	ug/L	12	Standard
>	**In**	115	649724.4	1.8				ug/L	587441	Standard
	Sn	118	890.0	5.4	**0.0430**	0.007	15.1	ug/L	497	Standard
	Sb	123	259622.2	1.1	**50.2896**	1.419	2.8	ug/L	78	Standard
	Ba	135	113320.2	1.2	**50.0586**	1.415	2.8	ug/L	24	Standard
	Ce	140	216.7	24.0				ug/L	25	Standard
>	**Tb**	159	1024515.0	4.0				ug/L	939701	Standard
	Ho	165	13.3	57.3				ug/L	10	Standard
	Tl	203	467714.3	0.2	**50.8953**	1.428	2.8	ug/L	4	Standard
	Tl	205	419104.4	1.0	**52.5470**	1.541	2.9	ug/L	2	Standard
	Pb	206	302725.5	0.9	**50.7682**	1.716	3.4	ug/L	240	Standard
	Pb	207	257785.2	0.8	**48.6571**	1.672	3.4	ug/L	198	Standard
	Pb	208	1092150.5	0.9	**52.1642**	0.990	1.9	ug/L	883	Standard
	U	238	389261.3	0.3	**50.7252**	1.507	3.0	ug/L	3	Standard
>	**Bi**	209	584427.6	2.7				ug/L	523984	Standard

Sample ID: LCSW 43 WG573291-04
Report Date/Time: Wednesday, June 29, 2016 20:22:54
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	56.7	20.4	**-0.0027**	0.015	535.8	mg/L	37	Standard
K	39	35.0	42.9	**-0.0491**	0.026	52.4	mg/L	38	Standard
Ca	43	18.3	15.7	**-1.5405**	0.154	10.0	mg/L	10	Standard
Fe	54	173.6	39.1	**-0.0145**	0.044	301.9	mg/L	179	Standard
Fe	57	231.7	7.6	**-0.4064**	0.037	9.2	mg/L	335	Standard
Sc-1	45	71925.5	3.7				mg/L	61029	Standard
Cl	35	25598.9	2.5				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1866.8	9.1				ug/L	1547	Standard
P	31	19263.9	1.4				ug/L	15286	Standard
S	34	2041.8	5.6				ug/L	1572	Standard
Sr	88	55.0	27.3				ug/L	73	Standard
C	12	76.7	37.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	19.2	93.9				mg/L	6	Standard
Ho-1	165	13.3	57.3				mg/L	10	Standard
Er	166	16.7	124.9				mg/L	13	Standard
I	127	1405.1	22.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		106.142	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		117.021	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: LCSW 43 WG573291-04
Report Date/Time: Wednesday, June 29, 2016 20:22:54
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.602
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.535
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 43 WG573291-04
Report Date/Time: Wednesday, June 29, 2016 20:22:54
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401 WG573291-01
Sample Date/Time: Wednesday, June 29, 2016 20:23:48
Number of Replicates: 3
Autosampler Position: 412
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81548.9	1.6				ug/L	74699	Standard
	Be	9	36.7	15.7	**0.0457**	0.006	12.1	ug/L	8	Standard
	Al	27	11596363.1	1.3	**72.2698**	0.564	0.8	ug/L	1493	Standard
	Sc	45	73884.7	1.3				ug/L	61029	Standard
	Ti	47	344.7	7.0	**0.7127**	0.059	8.2	ug/L	35	Standard
	V	51	2330.4	3.5	**0.0834**	0.015	17.8	ug/L	1435	Standard
	Cr	52	13306.8	2.8	**0.5995**	0.076	12.7	ug/L	7635	Standard
	Cr	53	1323.4	7.3	**0.9647**	0.110	11.4	ug/L	363	Standard
	Mn	55	1766896.9	0.7	**229.1909**	3.866	1.7	ug/L	747	Standard
	Co	59	11800.3	0.6	**1.5775**	0.021	1.4	ug/L	234	Standard
	Ni	60	3337.7	1.4	**1.4870**	0.026	1.7	ug/L	215	Standard
	Cu	65	729.0	2.5	**0.2502**	0.002	1.0	ug/L	104	Standard
	Zn	66	5843.1	1.1	**4.3791**	0.046	1.0	ug/L	363	Standard
>	**Ge**	72	548022.2	2.0				ug/L	495142	Standard
	As	75	368.2	8.1	**0.3569**	0.029	8.2	ug/L	-50	Standard
	Se	82	62.0	10.3	**0.3726**	0.060	16.1	ug/L	14	Standard
	Se-1	77	77.3	2.7	**0.3344**	0.023	6.7	ug/L	42	Standard
>	**Ga**	71	126.7	17.8				mg/L	42	Standard
	Rb	85	31524.4	1.4				ug/L	17	Standard
	Y	89	492847.4	3.2				ug/L	417098	Standard
>	**Rh**	103	60.0	14.4				ug/L	13	Standard
	Mo	98	169.7	8.5	**0.0488**	0.006	11.4	ug/L	11	Standard
	Ag	107	132.7	27.4	**0.0046**	0.005	112.7	ug/L	68	Standard
	Cd	111	39.0	30.5	**0.0081**	0.005	64.3	mg/L	2	Standard
	Cd	114	83.8	14.7	**0.0119**	0.002	19.5	ug/L	12	Standard
>	**In**	115	686408.4	2.0				ug/L	587441	Standard
	Sn	118	1350.1	6.2	**0.1033**	0.015	14.9	ug/L	497	Standard
	Sb	123	1853.5	25.3	**0.3348**	0.082	24.4	ug/L	78	Standard
	Ba	135	131659.9	0.9	**55.0560**	1.555	2.8	ug/L	24	Standard
	Ce	140	2260.2	0.8				ug/L	25	Standard
>	**Tb**	159	1046624.7	1.4				ug/L	939701	Standard
	Ho	165	75.0	29.1				ug/L	10	Standard
	Tl	203	216.3	9.4	**0.0226**	0.002	8.3	ug/L	4	Standard
	Tl	205	203.3	8.6	**0.0265**	0.002	7.2	ug/L	2	Standard
	Pb	206	535.3	4.4	**0.0217**	0.002	11.2	ug/L	240	Standard
	Pb	207	421.7	4.6	**0.0124**	0.003	25.0	ug/L	198	Standard
	Pb	208	1891.4	2.2	**0.0194**	0.001	4.1	ug/L	883	Standard
	U	238	92.7	15.8	**0.0080**	0.002	21.9	ug/L	3	Standard
>	**Bi**	209	565094.9	1.8				ug/L	523984	Standard

Sample ID: L1606097401 WG573291-01
Report Date/Time: Wednesday, June 29, 2016 20:26:05
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.0814**	27.845	173.2	mg/L	0	Standard
Mg	24	5874.5	4.9	**6.2016**	0.379	6.1	mg/L	37	Standard
K	39	696.7	5.3	**1.1574**	0.052	4.5	mg/L	38	Standard
Ca	43	108.3	19.2	**3.1206**	1.021	32.7	mg/L	10	Standard
Fe	54	752.2	10.4	**0.3885**	0.054	13.9	mg/L	179	Standard
Fe	57	410.0	11.6	**0.0774**	0.139	179.8	mg/L	335	Standard
Sc-1	45	73884.7	1.3				mg/L	61029	Standard
Cl	35	27603.8	2.8				ug/L	20195	Standard
Kr	83	3.0	66.7				ug/L	4	Standard
Br	81	11711.2	8.2				ug/L	1547	Standard
P	31	22032.7	3.4				ug/L	15286	Standard
S	34	2200.2	1.4				ug/L	1572	Standard
Sr	88	218.3	6.6				ug/L	73	Standard
C	12	103.3	43.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	96.2	10.7				mg/L	6	Standard
Ho-1	165	75.0	29.1				mg/L	10	Standard
Er	166	80.0	33.1				mg/L	13	Standard
I	127	173019.2	10.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		109.170	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		110.680	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606097401 WG573291-01
Report Date/Time: Wednesday, June 29, 2016 20:26:05
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.847
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.846
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097409S WG573291-05
Sample Date/Time: Wednesday, June 29, 2016 20:27:00
Number of Replicates: 3
Autosampler Position: 413
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	77620.1	3.1				ug/L	74699	Standard
	Be	9	53947.6	2.9	**58.9852**	0.884	1.5	ug/L	8	Standard
	Al	27	10156417.9	2.3	**66.5068**	0.592	0.9	ug/L	1493	Standard
	Sc	45	69663.2	3.3				ug/L	61029	Standard
	Ti	47	217.3	7.1	**0.4124**	0.032	7.7	ug/L	35	Standard
	V	51	458159.6	2.1	**55.0030**	1.779	3.2	ug/L	1435	Standard
	Cr	52	452840.8	0.8	**54.5717**	1.179	2.2	ug/L	7635	Standard
	Cr	53	56488.5	2.6	**55.2901**	2.221	4.0	ug/L	363	Standard
	Mn	55	2082250.4	1.8	**271.9842**	8.205	3.0	ug/L	747	Standard
	Co	59	416796.8	1.5	**57.3937**	1.632	2.8	ug/L	234	Standard
	Ni	60	112326.5	0.8	**54.4621**	1.176	2.2	ug/L	215	Standard
	Cu	65	110417.6	0.6	**53.8701**	0.759	1.4	ug/L	104	Standard
	Zn	66	70959.5	1.2	**58.6468**	1.068	1.8	ug/L	363	Standard
>	Ge	72	544304.5	1.4				ug/L	495142	Standard
	As	75	66583.4	2.1	**54.1492**	1.789	3.3	ug/L	-50	Standard
	Se	82	6605.0	1.3	**54.6200**	1.423	2.6	ug/L	14	Standard
	Se-1	77	4636.4	0.8	**57.1316**	0.988	1.7	ug/L	42	Standard
>	Ga	71	90.0	19.2				mg/L	42	Standard
	Rb	85	30654.3	2.7				ug/L	17	Standard
	Y	89	497057.5	2.9				ug/L	417098	Standard
>	Rh	103	38.3	52.7				ug/L	13	Standard
	Mo	98	163.0	7.4	**0.0487**	0.002	4.2	ug/L	11	Standard
	Ag	107	371099.4	2.0	**50.7567**	1.341	2.6	ug/L	68	Standard
	Cd	111	118377.9	2.8	**50.4636**	1.124	2.2	mg/L	2	Standard
	Cd	114	281427.9	1.4	**48.9324**	1.490	3.0	ug/L	12	Standard
>	In	115	659308.0	4.2				ug/L	587441	Standard
	Sn	118	1090.0	7.1	**0.0721**	0.019	25.7	ug/L	497	Standard
	Sb	123	259620.0	1.8	**49.5761**	1.212	2.4	ug/L	78	Standard
	Ba	135	232712.5	1.7	**101.3728**	3.182	3.1	ug/L	24	Standard
	Ce	140	941.7	8.0				ug/L	25	Standard
>	Tb	159	982549.7	3.5				ug/L	939701	Standard
	Ho	165	31.7	24.1				ug/L	10	Standard
	Tl	203	453507.3	1.5	**53.4353**	0.634	1.2	ug/L	4	Standard
	Tl	205	415041.8	2.5	**56.3402**	0.862	1.5	ug/L	2	Standard
	Pb	206	296774.5	1.1	**53.8916**	0.589	1.1	ug/L	240	Standard
	Pb	207	252251.9	1.7	**51.5506**	0.378	0.7	ug/L	198	Standard
	Pb	208	1071354.8	2.3	**55.4200**	1.065	1.9	ug/L	883	Standard
	U	238	381907.7	2.6	**53.8787**	0.717	1.3	ug/L	3	Standard
>	Bi	209	539522.1	2.0				ug/L	523984	Standard

Sample ID: L1606097409S WG573291-05
Report Date/Time: Wednesday, June 29, 2016 20:29:17
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.4920**	30.288	173.2	mg/L	0	Standard
Mg	24	5339.3	3.4	**5.9734**	0.106	1.8	mg/L	37	Standard
K	39	681.7	10.6	**1.2038**	0.100	8.3	mg/L	38	Standard
Ca	43	128.3	25.3	**4.6217**	2.038	44.1	mg/L	10	Standard
Fe	54	542.4	5.6	**0.2649**	0.028	10.4	mg/L	179	Standard
Fe	57	371.7	15.7	**0.0334**	0.176	526.0	mg/L	335	Standard
Sc-1	45	69663.2	3.3				mg/L	61029	Standard
Cl	35	28661.1	2.0				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	10680.4	5.1				ug/L	1547	Standard
P	31	21703.9	3.0				ug/L	15286	Standard
S	34	2123.5	4.7				ug/L	1572	Standard
Sr	88	153.3	22.2				ug/L	73	Standard
C	12	123.3	28.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	57.6	48.4				mg/L	6	Standard
Ho-1	165	31.7	24.1				mg/L	10	Standard
Er	166	50.0	91.7				mg/L	13	Standard
I	127	154285.8	7.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.911	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.929	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097409S WG573291-05
Report Date/Time: Wednesday, June 29, 2016 20:29:17
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.234
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.965
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606097409S WG573291-05
Report Date/Time: Wednesday, June 29, 2016 20:29:17
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097411SD WG573291-06
Sample Date/Time: Wednesday, June 29, 2016 20:30:11
Number of Replicates: 3
Autosampler Position: 414
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77405.6	1.5				ug/L	74699	Standard
	Be	9	55655.4	2.7	**61.0164**	1.576	2.6	ug/L	8	Standard
	Al	27	12086433.4	1.6	**79.3694**	1.985	2.5	ug/L	1493	Standard
	Sc	45	69212.8	2.6				ug/L	61029	Standard
	Ti	47	325.0	3.5	**0.6412**	0.031	4.8	ug/L	35	Standard
	V	51	466207.8	0.7	**53.8111**	1.021	1.9	ug/L	1435	Standard
	Cr	52	460631.7	1.4	**53.3535**	1.026	1.9	ug/L	7635	Standard
	Cr	53	56936.8	2.6	**53.5625**	1.349	2.5	ug/L	363	Standard
	Mn	55	2273187.5	1.8	**285.5286**	8.405	2.9	ug/L	747	Standard
	Co	59	417957.8	1.5	**55.3411**	1.481	2.7	ug/L	234	Standard
	Ni	60	114539.0	0.3	**53.3968**	0.715	1.3	ug/L	215	Standard
	Cu	65	111050.1	0.5	**52.0971**	0.842	1.6	ug/L	104	Standard
	Zn	66	69817.2	1.2	**55.4674**	1.299	2.3	ug/L	363	Standard
>	**Ge**	72	566018.8	1.2				ug/L	495142	Standard
	As	75	67402.1	1.8	**52.7097**	1.486	2.8	ug/L	-50	Standard
	Se	82	6620.3	2.0	**52.6418**	1.659	3.2	ug/L	14	Standard
	Se-1	77	4557.0	3.8	**53.9777**	2.700	5.0	ug/L	42	Standard
>	**Ga**	71	128.3	43.8				mg/L	42	Standard
	Rb	85	32409.6	0.5				ug/L	17	Standard
	Y	89	470382.0	1.0				ug/L	417098	Standard
>	**Rh**	103	61.7	24.8				ug/L	13	Standard
	Mo	98	196.1	5.9	**0.0618**	0.004	5.9	ug/L	11	Standard
	Ag	107	374581.1	0.3	**52.7275**	0.291	0.6	ug/L	68	Standard
	Cd	111	118658.6	1.5	**52.0656**	0.397	0.8	mg/L	2	Standard
	Cd	114	286274.0	2.6	**51.2138**	0.959	1.9	ug/L	12	Standard
>	**In**	115	640194.5	0.7				ug/L	587441	Standard
	Sn	118	1251.7	9.2	**0.1018**	0.017	16.3	ug/L	497	Standard
	Sb	123	263336.7	0.4	**51.7531**	0.327	0.6	ug/L	78	Standard
	Ba	135	250997.3	0.8	**112.5129**	0.934	0.8	ug/L	24	Standard
	Ce	140	2000.1	6.5				ug/L	25	Standard
>	**Tb**	159	984120.1	1.6				ug/L	939701	Standard
	Ho	165	48.3	11.9				ug/L	10	Standard
	Tl	203	459854.1	0.1	**53.6772**	0.823	1.5	ug/L	4	Standard
	Tl	205	414544.5	0.5	**55.7557**	1.089	2.0	ug/L	2	Standard
	Pb	206	298100.5	1.1	**53.6266**	1.109	2.1	ug/L	240	Standard
	Pb	207	253941.0	0.3	**51.4145**	0.875	1.7	ug/L	198	Standard
	Pb	208	1061848.2	0.5	**54.4146**	0.782	1.4	ug/L	883	Standard
	U	238	385197.1	0.6	**53.8449**	1.070	2.0	ug/L	3	Standard
>	**Bi**	209	544631.3	1.4				ug/L	523984	Standard

Sample ID: L1606097411SD WG573291-06
Report Date/Time: Wednesday, June 29, 2016 20:32:28
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	5929.5	2.1	**6.6847**	0.075	1.1	mg/L	37	Standard
K	39	801.7	4.2	**1.4479**	0.027	1.8	mg/L	38	Standard
Ca	43	125.0	31.2	**4.4423**	2.177	49.0	mg/L	10	Standard
Fe	54	841.7	6.4	**0.4916**	0.050	10.2	mg/L	179	Standard
Fe	57	441.7	5.4	**0.2489**	0.040	15.9	mg/L	335	Standard
> **Sc-1**	45	69212.8	2.6				mg/L	61029	Standard
Cl	35	28869.5	0.9				ug/L	20195	Standard
Kr	83	3.7	56.8				ug/L	4	Standard
Br	81	12511.8	0.5				ug/L	1547	Standard
P	31	22511.8	6.4				ug/L	15286	Standard
S	34	2320.2	4.3				ug/L	1572	Standard
Sr	88	170.0	8.8				ug/L	73	Standard
C	12	80.0	25.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	93.8	22.7				mg/L	6	Standard
Ho-1	165	48.3	11.9				mg/L	10	Standard
Er	166	60.0	57.7				mg/L	13	Standard
I	127	186712.7	7.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.624	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.314	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097411SD WG573291-06
Report Date/Time: Wednesday, June 29, 2016 20:32:28
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	108.980
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	103.940
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606097411SD WG573291-06
Report Date/Time: Wednesday, June 29, 2016 20:32:28
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401PS WG573790-03
Sample Date/Time: Wednesday, June 29, 2016 20:33:23
Number of Replicates: 3
Autosampler Position: 415
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	77767.4	1.4				ug/L	74699	Standard
	Be	9	54620.0	3.0	**59.5914**	1.002	1.7	ug/L	8	Standard
	Al	27	11452990.5	1.0	**74.8575**	1.594	2.1	ug/L	1493	Standard
	Sc	45	70666.1	1.7				ug/L	61029	Standard
	Ti	47	337.7	3.5	**0.6996**	0.031	4.4	ug/L	35	Standard
	V	51	448357.3	1.6	**53.6764**	0.056	0.1	ug/L	1435	Standard
	Cr	52	455460.2	1.4	**54.7521**	0.846	1.5	ug/L	7635	Standard
	Cr	53	57375.1	1.7	**56.0052**	0.200	0.4	ug/L	363	Standard
	Mn	55	2154709.7	0.5	**280.7294**	3.122	1.1	ug/L	747	Standard
	Co	59	404569.2	0.4	**55.5661**	0.653	1.2	ug/L	234	Standard
	Ni	60	112633.8	1.0	**54.4757**	0.921	1.7	ug/L	215	Standard
	Cu	65	109355.9	1.1	**53.2239**	1.011	1.9	ug/L	104	Standard
	Zn	66	70498.3	1.6	**58.1166**	0.789	1.4	ug/L	363	Standard
>	Ge	72	545627.2	1.6				ug/L	495142	Standard
	As	75	64850.8	0.9	**52.6029**	0.345	0.7	ug/L	-50	Standard
	Se	82	6485.5	1.8	**53.4847**	0.149	0.3	ug/L	14	Standard
	Se-1	77	4411.0	1.9	**54.1848**	0.787	1.5	ug/L	42	Standard
>	Ga	71	161.7	1.8				mg/L	42	Standard
	Rb	85	30524.1	1.7				ug/L	17	Standard
	Y	89	474875.4	0.7				ug/L	417098	Standard
>	Rh	103	60.0	16.7				ug/L	13	Standard
	Mo	98	163.5	8.5	**0.0491**	0.004	8.3	ug/L	11	Standard
	Ag	107	365805.1	1.5	**50.1795**	1.312	2.6	ug/L	68	Standard
	Cd	111	117534.8	0.9	**50.2525**	0.249	0.5	mg/L	2	Standard
	Cd	114	285376.1	1.5	**49.7475**	0.497	1.0	ug/L	12	Standard
>	In	115	657058.7	1.4				ug/L	587441	Standard
	Sn	118	1523.4	8.4	**0.1383**	0.017	12.1	ug/L	497	Standard
	Sb	123	259036.8	1.0	**49.6101**	1.139	2.3	ug/L	78	Standard
	Ba	135	239792.4	1.1	**104.7386**	1.629	1.6	ug/L	24	Standard
	Ce	140	2208.5	2.2				ug/L	25	Standard
>	Tb	159	1019758.9	1.4				ug/L	939701	Standard
	Ho	165	71.7	21.3				ug/L	10	Standard
	Tl	203	456925.9	1.0	**51.8248**	0.434	0.8	ug/L	4	Standard
	Tl	205	410087.8	1.3	**53.5969**	1.082	2.0	ug/L	2	Standard
	Pb	206	289464.5	1.3	**50.5998**	1.170	2.3	ug/L	240	Standard
	Pb	207	248071.4	0.9	**48.8073**	1.130	2.3	ug/L	198	Standard
	Pb	208	1050860.8	0.8	**52.3325**	1.144	2.2	ug/L	883	Standard
	U	238	380364.0	0.7	**51.6678**	1.019	2.0	ug/L	3	Standard
>	Bi	209	560465.1	1.6				ug/L	523984	Standard

Sample ID: L1606097401PS WG573790-03
Report Date/Time: Wednesday, June 29, 2016 20:35:40
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.2901**	28.207	173.2	mg/L	0	Standard
Mg	24	5812.8	3.2	**6.4153**	0.191	3.0	mg/L	37	Standard
K	39	706.7	5.8	**1.2347**	0.072	5.8	mg/L	38	Standard
Ca	43	108.3	32.4	**3.3848**	1.922	56.8	mg/L	10	Standard
Fe	54	723.0	16.6	**0.3920**	0.096	24.5	mg/L	179	Standard
Fe	57	435.0	5.7	**0.2027**	0.073	35.8	mg/L	335	Standard
Sc-1	45	70666.1	1.7				mg/L	61029	Standard
Cl	35	28625.7	0.6				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	11964.7	4.0				ug/L	1547	Standard
P	31	22167.9	1.3				ug/L	15286	Standard
S	34	2268.5	2.8				ug/L	1572	Standard
Sr	88	201.7	11.5				ug/L	73	Standard
C	12	93.3	34.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	3	Standard
Dy	164	120.1	22.8				mg/L	6	Standard
Ho-1	165	71.7	21.3				mg/L	10	Standard
Er	166	66.7	60.6				mg/L	13	Standard
I	127	179600.9	7.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		104.108	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		110.196	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606097401PS WG573790-03
Report Date/Time: Wednesday, June 29, 2016 20:35:40
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.851
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.962
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606097401PS WG573790-03
Report Date/Time: Wednesday, June 29, 2016 20:35:40
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401SDL WG573790-04
Sample Date/Time: Wednesday, June 29, 2016 20:36:34
Number of Replicates: 3
Autosampler Position: 416
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77070.7	3.6				ug/L	74699	Standard
	Be	9	45.0	33.3	**0.0568**	0.015	26.1	ug/L	8	Standard
	Al	27	2260304.8	1.2	**14.9037**	0.448	3.0	ug/L	1493	Standard
	Sc	45	68278.6	3.1				ug/L	61029	Standard
	Ti	47	83.0	9.9	**0.0892**	0.019	21.8	ug/L	35	Standard
	V	51	2192.5	3.4	**0.0673**	0.008	12.3	ug/L	1435	Standard
	Cr	52	10144.1	1.9	**0.2156**	0.020	9.5	ug/L	7635	Standard
	Cr	53	686.7	4.8	**0.3429**	0.034	10.1	ug/L	363	Standard
	Mn	55	355711.8	1.3	**46.1130**	0.456	1.0	ug/L	747	Standard
	Co	59	2492.2	1.8	**0.3045**	0.007	2.2	ug/L	234	Standard
	Ni	60	828.7	2.9	**0.2769**	0.011	3.8	ug/L	215	Standard
	Cu	65	280.0	0.7	**0.0325**	0.001	2.6	ug/L	104	Standard
	Zn	66	2517.5	2.7	**1.6325**	0.059	3.6	ug/L	363	Standard
>	**Ge**	72	546825.8	0.3				ug/L	495142	Standard
	As	75	65.5	45.7	**0.1126**	0.024	21.6	ug/L	-50	Standard
	Se	82	28.6	15.9	**0.0977**	0.038	38.5	ug/L	14	Standard
	Se-1	77	61.3	12.0	**0.1380**	0.090	65.2	ug/L	42	Standard
>	**Ga**	71	98.3	5.9				mg/L	42	Standard
	Rb	85	6301.3	2.8				ug/L	17	Standard
	Y	89	489576.3	2.2				ug/L	417098	Standard
>	**Rh**	103	21.7	48.0				ug/L	13	Standard
	Mo	98	47.0	21.8	**0.0100**	0.004	35.1	ug/L	11	Standard
	Ag	107	179.0	15.1	**0.0115**	0.004	30.5	ug/L	68	Standard
	Cd	111	35.6	11.4	**0.0072**	0.002	22.2	mg/L	2	Standard
	Cd	114	102.0	28.8	**0.0155**	0.005	31.8	ug/L	12	Standard
>	**In**	115	661161.0	0.8				ug/L	587441	Standard
	Sn	118	795.0	15.7	**0.0261**	0.018	68.8	ug/L	497	Standard
	Sb	123	2322.3	23.7	**0.4382**	0.107	24.4	ug/L	78	Standard
	Ba	135	26230.7	1.3	**11.3705**	0.051	0.4	ug/L	24	Standard
	Ce	140	480.0	13.3				ug/L	25	Standard
>	**Tb**	159	1000122.3	1.4				ug/L	939701	Standard
	Ho	165	26.7	10.8				ug/L	10	Standard
	Tl	203	156.0	22.9	**0.0163**	0.004	24.4	ug/L	4	Standard
	Tl	205	150.0	15.3	**0.0202**	0.003	14.3	ug/L	2	Standard
	Pb	206	403.0	8.6	**0.0007**	0.006	871.4	ug/L	240	Standard
	Pb	207	342.3	5.9	**-0.0013**	0.004	288.2	ug/L	198	Standard
	Pb	208	1497.0	5.1	**0.0020**	0.004	181.5	ug/L	883	Standard
	U	238	114.0	13.2	**0.0113**	0.002	17.4	ug/L	3	Standard
>	**Bi**	209	549941.1	1.0				ug/L	523984	Standard

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:38:51
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	1213.4	3.4	**1.3365**	0.084	6.3	mg/L	37	Standard
K	39	180.0	11.1	**0.2425**	0.050	20.5	mg/L	38	Standard
Ca	43	46.7	22.3	**0.1064**	0.547	514.2	mg/L	10	Standard
Fe	54	286.0	10.0	**0.0782**	0.017	21.6	mg/L	179	Standard
Fe	57	306.7	10.9	**-0.1437**	0.079	55.0	mg/L	335	Standard
> **Sc-1**	45	68278.6	3.1				mg/L	61029	Standard
Cl	35	28506.1	0.3				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	3947.2	6.9				ug/L	1547	Standard
P	31	19479.2	3.3				ug/L	15286	Standard
S	34	2088.5	5.3				ug/L	1572	Standard
Sr	88	101.7	24.8				ug/L	73	Standard
C	12	60.0	60.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	52.5	9.9				mg/L	6	Standard
Ho-1	165	26.7	10.8				mg/L	10	Standard
Er	166	16.7	69.3				mg/L	13	Standard
I	127	39316.0	2.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.175	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.438	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:38:51
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.549
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.954
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401SDL WG573790-04
Sample Date/Time: Wednesday, June 29, 2016 20:39:46
Number of Replicates: 3
Autosampler Position: 417
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76547.9	2.1				ug/L	74699	Standard
	Be	9	28.3	27.0	**0.0389**	0.008	20.5	ug/L	8	Standard
	Al	27	465099.1	0.7	**3.0756**	0.044	1.4	ug/L	1493	Standard
	Sc	45	68169.7	0.9				ug/L	61029	Standard
	Ti	47	36.7	6.9	**-0.0243**	0.006	24.2	ug/L	35	Standard
	V	51	2063.3	3.2	**0.0438**	0.007	16.8	ug/L	1435	Standard
	Cr	52	9662.1	0.8	**0.1188**	0.007	6.3	ug/L	7635	Standard
	Cr	53	523.3	10.2	**0.1664**	0.051	30.7	ug/L	363	Standard
	Mn	55	72938.2	2.1	**9.0325**	0.175	1.9	ug/L	747	Standard
	Co	59	731.7	15.3	**0.0598**	0.015	24.7	ug/L	234	Standard
	Ni	60	309.0	4.2	**0.0207**	0.006	30.5	ug/L	215	Standard
	Cu	65	310.7	4.3	**0.0426**	0.006	14.7	ug/L	104	Standard
	Zn	66	2601.2	2.1	**1.6324**	0.044	2.7	ug/L	363	Standard
>	**Ge**	72	565002.4	0.2				ug/L	495142	Standard
	As	75	-33.8	92.8	**0.0331**	0.025	74.3	ug/L	-50	Standard
	Se	82	20.9	22.0	**0.0285**	0.036	128.0	ug/L	14	Standard
	Se-1	77	57.7	17.5	**0.0697**	0.122	174.5	ug/L	42	Standard
>	**Ga**	71	58.3	9.9				mg/L	42	Standard
	Rb	85	1335.1	7.2				ug/L	17	Standard
	Y	89	490772.8	1.5				ug/L	417098	Standard
>	**Rh**	103	15.0	57.7				ug/L	13	Standard
	Mo	98	24.2	10.4	**0.0025**	0.001	32.4	ug/L	11	Standard
	Ag	107	143.3	39.3	**0.0071**	0.008	112.3	ug/L	68	Standard
	Cd	111	11.6	65.7	**-0.0029**	0.003	116.5	mg/L	2	Standard
	Cd	114	25.4	61.7	**0.0024**	0.003	115.9	ug/L	12	Standard
>	**In**	115	647709.2	0.5				ug/L	587441	Standard
	Sn	118	543.3	5.3	**-0.0103**	0.004	42.8	ug/L	497	Standard
	Sb	123	784.5	24.6	**0.1481**	0.037	24.8	ug/L	78	Standard
	Ba	135	5327.9	0.7	**2.3449**	0.015	0.6	ug/L	24	Standard
	Ce	140	110.0	7.9				ug/L	25	Standard
>	**Tb**	159	986009.2	0.5				ug/L	939701	Standard
	Ho	165	10.0	100.0				ug/L	10	Standard
	Tl	203	50.7	34.9	**0.0041**	0.002	51.6	ug/L	4	Standard
	Tl	205	46.7	71.3	**0.0064**	0.005	70.3	ug/L	2	Standard
	Pb	206	376.3	19.7	**-0.0043**	0.014	323.1	ug/L	240	Standard
	Pb	207	290.7	20.1	**-0.0119**	0.012	103.4	ug/L	198	Standard
	Pb	208	1284.0	11.0	**-0.0091**	0.008	86.4	ug/L	883	Standard
	U	238	44.0	79.9	**0.0017**	0.005	294.8	ug/L	3	Standard
>	**Bi**	209	552760.9	1.2				ug/L	523984	Standard

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:42:03
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	246.7	9.1	**0.2201**	0.028	12.6	mg/L	37	Standard
K	39	50.0	40.0	**-0.0154**	0.039	250.4	mg/L	38	Standard
Ca	43	20.0	43.3	**-1.3952**	0.477	34.2	mg/L	10	Standard
Fe	54	182.5	25.5	**0.0004**	0.036	9653.1	mg/L	179	Standard
Fe	57	256.7	8.8	**-0.2933**	0.068	23.1	mg/L	335	Standard
Sc-1	45	68169.7	0.9				mg/L	61029	Standard
Cl	35	28987.0	1.4				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2360.2	9.1				ug/L	1547	Standard
P	31	19090.4	2.0				ug/L	15286	Standard
S	34	2190.2	0.6				ug/L	1572	Standard
Sr	88	73.3	27.6				ug/L	73	Standard
C	12	46.7	44.6				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	22.7	102.4				mg/L	6	Standard
Ho-1	165	10.0	100.0				mg/L	10	Standard
Er	166	13.3	114.6				mg/L	13	Standard
I	127	10453.6	7.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.475	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.109	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:42:03
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.259
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.492
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 20:42:59
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	73968.6	3.7				ug/L	74699	Standard
	Be	9	51786.8	4.7	**59.3953**	0.618	1.0	ug/L	8	Standard
	Al	27	8138418.2	2.1	**55.9353**	1.016	1.8	ug/L	1493	Standard
	Sc	45	66422.0	2.7				ug/L	61029	Standard
	Ti	47	44367.5	1.4	**105.9251**	2.261	2.1	ug/L	35	Standard
	V	51	448161.7	0.9	**53.5504**	0.563	1.1	ug/L	1435	Standard
	Cr	52	438098.2	0.5	**52.5206**	0.674	1.3	ug/L	7635	Standard
	Cr	53	53648.2	1.8	**52.2574**	1.658	3.2	ug/L	363	Standard
	Mn	55	415062.2	0.8	**53.8560**	1.244	2.3	ug/L	747	Standard
	Co	59	407202.7	0.3	**55.8201**	1.033	1.9	ug/L	234	Standard
	Ni	60	107214.7	0.8	**51.7442**	0.748	1.4	ug/L	215	Standard
	Cu	65	106835.8	1.9	**51.8884**	1.090	2.1	ug/L	104	Standard
	Zn	66	63121.9	0.7	**51.8882**	0.831	1.6	ug/L	363	Standard
>	Ge	72	546729.2	1.7				ug/L	495142	Standard
	As	75	63950.8	1.3	**51.7792**	1.311	2.5	ug/L	-50	Standard
	Se	82	6522.2	2.4	**53.7001**	1.978	3.7	ug/L	14	Standard
	Se-1	77	4462.7	1.0	**54.7224**	0.944	1.7	ug/L	42	Standard
>	Ga	71	48.3	11.9				mg/L	42	Standard
	Rb	85	921.7	10.0				ug/L	17	Standard
	Y	89	469877.0	1.2				ug/L	417098	Standard
>	Rh	103	31.7	45.6				ug/L	13	Standard
	Mo	98	306064.1	0.6	**105.7488**	1.650	1.6	ug/L	11	Standard
	Ag	107	358011.4	0.7	**50.5757**	0.461	0.9	ug/L	68	Standard
	Cd	111	116953.2	1.0	**51.5035**	0.389	0.8	mg/L	2	Standard
	Cd	114	288128.0	0.1	**51.7395**	0.785	1.5	ug/L	12	Standard
>	In	115	637946.0	1.6				ug/L	587441	Standard
	Sn	118	327214.8	2.0	**51.4831**	0.238	0.5	ug/L	497	Standard
	Sb	123	254525.5	0.5	**50.2034**	0.722	1.4	ug/L	78	Standard
	Ba	135	108828.5	1.0	**48.9489**	0.332	0.7	ug/L	24	Standard
	Ce	140	90.0	29.4				ug/L	25	Standard
>	Tb	159	958314.6	1.8				ug/L	939701	Standard
	Ho	165	11.7	49.5				ug/L	10	Standard
	Tl	203	447416.0	0.6	**53.4811**	1.019	1.9	ug/L	4	Standard
	Tl	205	409322.9	1.6	**56.3810**	1.712	3.0	ug/L	2	Standard
	Pb	206	283417.6	0.7	**52.2045**	0.689	1.3	ug/L	240	Standard
	Pb	207	252263.1	0.7	**52.2984**	0.513	1.0	ug/L	198	Standard
	Pb	208	1062273.6	1.1	**55.7538**	1.546	2.8	ug/L	883	Standard
	U	238	385444.7	0.9	**55.1725**	1.082	2.0	ug/L	3	Standard
>	Bi	209	531871.1	1.6				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4377.3	2.2	**5.1285**	0.149	2.9	mg/L	37	Standard
K	39	2180.2	1.8	**4.3148**	0.109	2.5	mg/L	38	Standard
Ca	43	118.3	10.6	**4.3366**	0.588	13.6	mg/L	10	Standard
Fe	54	6990.5	4.3	**5.3088**	0.233	4.4	mg/L	179	Standard
Fe	57	2083.5	2.8	**5.4386**	0.352	6.5	mg/L	335	Standard
> **Sc-1**	45	66422.0	2.7				mg/L	61029	Standard
Cl	35	29309.7	0.2				ug/L	20195	Standard
Kr	83	2.3	24.7				ug/L	4	Standard
Br	81	2156.8	3.9				ug/L	1547	Standard
P	31	21223.2	2.6				ug/L	15286	Standard
S	34	2665.2	2.0				ug/L	1572	Standard
Sr	88	71.7	8.1				ug/L	73	Standard
C	12	56.7	40.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	19.8	1.4				mg/L	6	Standard
Ho-1	165	11.7	49.5				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	2623.6	23.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	118.791		
Al	27	111.871		
Sc	45			
Ti	47	105.925		
V	51	107.101		
Cr	52	105.041		
Cr	53			
Mn	55	107.712		
Co	59	111.640		
Ni	60	103.488		
Cu	65	103.777		
Zn	66	103.776		
> Ge	72		110.419	
As	75	103.558		
Se	82	107.400		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	105.749		
Ag	107	101.151		
Cd	111	103.007		
Cd	114			
> In	115		108.597	
Sn	118	102.966		
Sb	123	100.407		
Ba	135	97.898		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	106.962		
Tl	205			
Pb	206			
Pb	207			
Pb	208	111.508		
U	238	110.345		
> Bi	209		101.505	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 3

Approved: July 04, 2016

QC Std 6	Pb	208
QC Std 6	U	238

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 20:46:11
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75383.8	1.3				ug/L	74699	Standard
	Be	9	28.3	20.4	**0.0395**	0.006	15.9	ug/L	8	Standard
	Al	27	2638.6	20.0	**0.0044**	0.003	78.6	ug/L	1493	Standard
	Sc	45	68713.9	3.5				ug/L	61029	Standard
	Ti	47	37.7	13.4	**-0.0232**	0.011	49.4	ug/L	35	Standard
	V	51	1533.2	3.6	**-0.0202**	0.008	41.3	ug/L	1435	Standard
	Cr	52	7879.4	2.2	**-0.1054**	0.029	27.9	ug/L	7635	Standard
	Cr	53	448.3	16.3	**0.0885**	0.063	71.1	ug/L	363	Standard
	Mn	55	901.4	5.8	**-0.0363**	0.007	18.9	ug/L	747	Standard
	Co	59	283.3	10.0	**-0.0002**	0.004	1542.3	ug/L	234	Standard
	Ni	60	186.3	2.6	**-0.0379**	0.003	8.3	ug/L	215	Standard
	Cu	65	132.7	3.4	**-0.0421**	0.001	3.4	ug/L	104	Standard
	Zn	66	415.3	5.4	**-0.1263**	0.014	10.9	ug/L	363	Standard
>	**Ge**	72	573402.3	1.2				ug/L	495142	Standard
	As	75	-34.2	13.3	**0.0332**	0.003	10.5	ug/L	-50	Standard
	Se	82	17.2	23.7	**-0.0025**	0.031	1279.2	ug/L	14	Standard
	Se-1	77	54.3	12.3	**0.0204**	0.082	402.5	ug/L	42	Standard
>	**Ga**	71	35.0	37.8				mg/L	42	Standard
	Rb	85	30.0	28.9				ug/L	17	Standard
	Y	89	465691.4	1.8				ug/L	417098	Standard
>	**Rh**	103	6.7	43.3				ug/L	13	Standard
	Mo	98	268.8	25.2	**0.0874**	0.025	29.0	ug/L	11	Standard
	Ag	107	132.7	5.0	**0.0058**	0.001	22.7	ug/L	68	Standard
	Cd	111	15.2	35.4	**-0.0012**	0.003	209.6	mg/L	2	Standard
	Cd	114	58.4	20.0	**0.0084**	0.002	23.7	ug/L	12	Standard
>	**In**	115	638175.3	2.2				ug/L	587441	Standard
	Sn	118	866.7	15.8	**0.0416**	0.019	45.6	ug/L	497	Standard
	Sb	123	616.5	30.1	**0.1169**	0.034	29.1	ug/L	78	Standard
	Ba	135	50.7	18.3	**0.0068**	0.005	67.0	ug/L	24	Standard
	Ce	140	28.3	27.0				ug/L	25	Standard
>	**Tb**	159	990005.2	3.0				ug/L	939701	Standard
	Ho	165	6.7	86.6				ug/L	10	Standard
	Tl	203	92.0	24.5	**0.0086**	0.003	32.5	ug/L	4	Standard
	Tl	205	76.7	43.4	**0.0102**	0.005	44.4	ug/L	2	Standard
	Pb	206	305.7	10.6	**-0.0179**	0.007	38.8	ug/L	240	Standard
	Pb	207	274.7	5.9	**-0.0162**	0.004	23.3	ug/L	198	Standard
	Pb	208	1058.7	7.3	**-0.0216**	0.005	23.5	ug/L	883	Standard
	U	238	60.3	25.7	**0.0037**	0.002	61.2	ug/L	3	Standard
>	**Bi**	209	564302.2	2.4				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:48:28
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	35.0	37.8	**-0.0253**	0.014	54.1	mg/L	37	Standard
K	39	45.0	19.2	**-0.0261**	0.014	54.4	mg/L	38	Standard
Ca	43	26.7	39.0	**-1.0220**	0.608	59.5	mg/L	10	Standard
Fe	54	162.5	17.8	**-0.0163**	0.018	110.8	mg/L	179	Standard
Fe	57	258.3	8.1	**-0.2948**	0.047	15.9	mg/L	335	Standard
Sc-1	45	68713.9	3.5				mg/L	61029	Standard
Cl	35	28935.6	0.2				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	1740.1	12.2				ug/L	1547	Standard
P	31	21446.8	2.8				ug/L	15286	Standard
S	34	2463.5	5.1				ug/L	1572	Standard
Sr	88	71.7	17.6				ug/L	73	Standard
C	12	43.3	13.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	12.9	115.2				mg/L	6	Standard
Ho-1	165	6.7	86.6				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	2023.5	28.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.806	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:48:28
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.636
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.694
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:48:28
Page 3

Approved: July 04, 2016

2.3.3 Metals CVAA Data (Mercury)

2.3.3.1 Summary Data

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-02	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	126F-061416	**Prep Method:**	7470A	**Prep Date:**	06/16/2016 07:27
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/26/2016 12:23
Workgroup #:	WG573236	**Analyst:**	AMA	**Run Date:**	06/26/2016 13:19
Collect Date:	06/14/2016 11:10	**Dilution:**	1	**File ID:**	M7.062616.131930
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-03	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	C08-061416	**Prep Method:**	7470A	**Prep Date:**	06/16/2016 07:27
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/26/2016 12:23
Workgroup #:	WG573236	**Analyst:**	AMA	**Run Date:**	06/26/2016 13:22
Collect Date:	06/14/2016 13:10	**Dilution:**	1	**File ID:**	M7.062616.132202
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060823-04	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	C03-061416	**Prep Method:**	7470A	**Prep Date:**	06/16/2016 07:27
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/26/2016 12:23
Workgroup #:	WG573236	**Analyst:**	AMA	**Run Date:**	06/26/2016 13:24
Collect Date:	06/14/2016 14:15	**Dilution:**	1	**File ID:**	M7.062616.132434
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060823
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.3.3.2 QC Summary

Example Cold Vapor Mercury Calculations
Hydra AA Mercury Analyzer / CETAC M-7600 Quick Trace Mercury Analyzer

1.0 Initial Calibration (ICAL) Parameters
The system performs linear regression from data consisting of a blank and five standards.

2.0 Calculating the concentration (C) of an element in water using data from run log and quantitation report (note:the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (ug/L)	0.1
Vf = Diluted to Volume (mL)	40
Vi = Aliquot Volume (mL)	40
D = Manual dilution factor, if required (10X = 10)	1
Cx = Concentration of element in ppb (ug/L)	0.1

3.0 Calculating the concentration (C) of an element in soil using data from prep log and quantitation report (note: the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Ws} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (ug/L)	0.1
Vf = Diluted to volume (mL)	40
Ws = Aliquot weight (g)	0.6
D = Manual dilution factor	1
Cx = Concentration of element in ug/kg	6.67

4.0 Adjusting the concentration to dry weight:

$$Cdry = \frac{Cx \times 100}{Px}$$

1 Cx = Concentration calculated as received (wet basis)	6.67
Px = Percent solids of sample (%wt)	80
$Cdry$ = Concentration calculated as dry weight (ug/kg)	8.33

8.33 ug/kg = 0.00833 mg/kg

Microbac Laboratories Inc.
Metals Digest Log

Workgroup: WG572915
Analyst: REK
Spike Analyst: REK
Method: 7470A
Run Date: 06/16/2016 07:27
Hotblock Start Temp: 95.2 @ 07:15
Hotblock End Temp: 95 @ 09:15
Instrument: HB7

SOP: ME404 Revison 17
Spike Solution: STD76639
Spike Witness: VC
H2SO4 Lot #: COA18359
40 & 50 ML. DIGESTION TU COA18772
HNO3 Lot #: COA18905
K2S2O8 1:1 Lot #: RGT37046
KMnO4 1:1 Lot #: RGT37082
Mercury Water ICV Lot #: STD76641
HG H2O STDS 10PPM Lot #: STD76647

	SAMPLE #	Type	Matrix	Initial Amount	Final Volume	Spike Amount	Due Date
1	WG572915-02	BLANK	1	40 mL	40 mL		
2	WG572915-03	LCS	1	40 mL	40 mL	4 mL	
3	L16060764-01	SAMP	1	40 mL	40 mL		06/28/16
4	L16060764-02	SAMP	1	40 mL	40 mL		06/28/16
5	L16060764-03	SAMP	1	40 mL	40 mL		06/28/16
6	WG572915-01	REF	1	40 mL	40 mL		
7	L16060764-04	RS01	1	40 mL	40 mL		06/28/16
8	WG572915-04	MS	1	36 mL	40 mL	4 mL	
9	L16060764-05	MS01	1	36 mL	40 mL	4 mL	06/28/16
10	WG572915-05	MSD	1	36 mL	40 mL	4 mL	
11	L16060764-06	SD01	1	36 mL	40 mL	4 mL	06/28/16
12	L16060794-01	SAMP	1	40 mL	40 mL		06/24/16
13	L16060794-02	SAMP	1	40 mL	40 mL		06/24/16
14	L16060794-03	SAMP	1	40 mL	40 mL		06/24/16
15	L16060794-04	SAMP	1	40 mL	40 mL		06/24/16
16	L16060822-02	SAMP	1	40 mL	40 mL		06/24/16
17	L16060822-03	SAMP	1	40 mL	40 mL		06/24/16
18	L16060823-02	SAMP	1	40 mL	40 mL		06/24/16
19	L16060823-03	SAMP	1	40 mL	40 mL		06/24/16
20	L16060823-04	SAMP	1	40 mL	40 mL		06/24/16
21	L16060840-01	SAMP	1	40 mL	40 mL		06/29/16
22	L16060840-02	SAMP	1	40 mL	40 mL		06/29/16
23	L16060840-03	SAMP	1	40 mL	40 mL		06/29/16
24	L16060840-04	SAMP	1	40 mL	40 mL		06/29/16
25	L16060840-05	SAMP	1	40 mL	40 mL		06/29/16
26	L16060840-06	SAMP	1	40 mL	40 mL		06/29/16

Analyst: [signature]

Reviewer: [signature]

* All calibration and check standards are prepared and digested with sample batch following the procedures in section 7.0 of SOP ME404/ME405.

HB_DIG - Modified 07/26/2012
PDF ID: 4809801
Report generated: 06/16/2016 12:16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: CVAA1 Dataset: 062616A.CSV
Analyst1: AMA Analyst2: N/A
Method: 7470A/245.1 SOP: ME404 Rev: 17
Maintenance Log ID:

Calibration Std: STD76647 ICV Std: STD76641 Post Spike: STD76647
ICSA: N/A ICSAB: N/A Int. Std:
CCV: LLCCV: Tuning Sol :
Stannous : RGT36627 Hydroxylamine : RGT36625

Workgroups: 573236

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	M7.062616.121106	WG574322-01	Calibration Point		1		06/26/16 12:11
2	M7.062616.121337	WG574322-02	Calibration Point		1		06/26/16 12:13
3	M7.062616.121609	WG574322-03	Calibration Point		1		06/26/16 12:16
4	M7.062616.121841	WG574322-04	Calibration Point		1		06/26/16 12:18
5	M7.062616.122114	WG574322-05	Calibration Point		1		06/26/16 12:21
6	M7.062616.122348	WG574322-06	Calibration Point		1		06/26/16 12:23
7	M7.062616.122622	WG574322-07	Initial Calibration Verification		1		06/26/16 12:26
8	M7.062616.122853	WG574322-08	Initial Calib Blank		1		06/26/16 12:28
9	M7.062616.123125	WG574322-09	CCV		1		06/26/16 12:31
10	M7.062616.123356	WG574322-10	CCB		1		06/26/16 12:33
11	M7.062616.123627	WG572915-02	Method/Prep Blank	40/40	1		06/26/16 12:36
12	M7.062616.123858	WG572915-03	Laboratory Control S	40/40	1		06/26/16 12:38
13	M7.062616.124129	L16060764-01	MW-23-004	40/40	1		06/26/16 12:41
14	M7.062616.124401	WG573236-01	Post Digestion Spike		1	L16060764-01	06/26/16 12:44
15	M7.062616.124633	L16060764-02	MW-23-002		1		06/26/16 12:46
16	M7.062616.124905	L16060764-03	FD-061316		1		06/26/16 12:49
17	M7.062616.125137	WG572915-05	Matrix Spike Duplica	36/40	1	L16060764-04	06/26/16 12:51
18	M7.062616.125410	L16060764-04	MW-23-005		1	WG572915-01	06/26/16 12:54
19	M7.062616.125643	WG572915-04	Matrix Spike	36/40	1	L16060764-04	06/26/16 12:56
20	M7.062616.125916	L16060794-01	A08-MW06-Y1S2	40/40	1		06/26/16 12:59
21	M7.062616.130149	WG574322-11	CCV		1		06/26/16 13:01
22	M7.062616.130420	WG574322-12	CCB		1		06/26/16 13:04
23	M7.062616.130654	L16060794-02	A08-MW09-Y1S2	40/40	1		06/26/16 13:06
24	M7.062616.130927	L16060794-03	A09-MW02-Y1S2	40/40	1		06/26/16 13:09
25	M7.062616.131158	L16060794-04	A09-MW07-Y1S2	40/40	1		06/26/16 13:11
26	M7.062616.131428	L16060822-02	18WW03F-061416	40/40	1		06/26/16 13:14
27	M7.062616.131659	L16060822-03	18CPTMW04SW-061416	40/40	1		06/26/16 13:16
28	M7.062616.131930	L16060823-02	126F-061416	40/40	1		06/26/16 13:19
29	M7.062616.132202	L16060823-03	C08-061416	40/40	1		06/26/16 13:22
30	M7.062616.132434	L16060823-04	C03-061416	40/40	1		06/26/16 13:24
31	M7.062616.132706	L16060840-01	MW-24-004R	40/40	1		06/26/16 13:27
32	M7.062616.132938	L16060840-02	MW-24-003R	40/40	1		06/26/16 13:29
33	M7.062616.133211	WG574322-13	CCV		1		06/26/16 13:32
34	M7.062616.133441	WG574322-14	CCB		1		06/26/16 13:34

Page: 1 Approved: June 29, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: CVAA1 Dataset: 062616A.CSV

Analyst1: AMA Analyst2: N/A

Method: 7470A/245.1 SOP: ME404 Rev: 17

Maintenance Log ID:

Calibration Std: STD76647 ICV Std: STD76641 Post Spike: STD76647

ICSA: N/A ICSAB: N/A Int. Std:

CCV: LLCCV: Tuning Sol :

Stannous : RGT36627 Hydroxylamine : RGT36625

Workgroups: 573236

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	M7.062616.133714	L16060840-03	FR-061416-24	40/40	1		06/26/16 13:37
36	M7.062616.133947	L16060840-04	MW-24-002R	40/40	1		06/26/16 13:39
37	M7.062616.134220	L16060840-05	MW-24-005	40/40	1		06/26/16 13:42
38	M7.062616.134454	L16060840-06	EB-LOW FLOW TUBING	40/40	1		06/26/16 13:44
39	M7.062616.134726	WG574322-15	CCV		1		06/26/16 13:47
40	M7.062616.134957	WG574322-16	CCB		1		06/26/16 13:49

Page: 2 Approved: June 29, 2016

Microbac Laboratories Inc.

Data Checklist

Date: 26-JUN-2016
Analyst: AMA
Analyst: NA
Method: 7470A/245.1
Instrument: CVAA1
Curve Workgroup: 574322
Runlog ID: 75946
Analytical Workgroups: 573236

STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	
ICB/CCB	X
ICSA/ICSAB	
CRI	
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	0794, 0822, 0823, 0840
Client Forms	X
Level X	
Level 3	0840
Level 4	0794, 0822, 0823
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	AMA
Secondary Reviewer	BKT
Comments	

Primary Reviewer:
28-JUN-2016
Alexandra M. Alfred

Secondary Reviewer:
29-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-29-2016 12:56:19

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 7470A

AAB#: WG573236

Login Number: L16060823

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
126F-061416	02	06/14/16					06/16/2016	1.8	28		06/26/16	12.1	28	
C08-061416	03	06/14/16					06/16/2016	1.8	28		06/26/16	12	28	
C03-061416	04	06/14/16					06/16/2016	1.7	28		06/26/16	12	28	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4828279
Report generated 06/28/2016 10:48

METHOD BLANK SUMMARY

Login Number: L16060823
Blank File ID: M7.062616.123627
Prep Date: 06/16/16 07:27
Analyzed Date: 06/26/16 12:36
Analyst: AMA

Work Group: WG573236
Blank Sample ID: WG572915-02
Instrument ID: CVAA1
Method: 7470A

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG572915-03	M7.062616.123858	06/26/16 12:38	01
126F-061416	L16060823-02	M7.062616.131930	06/26/16 13:19	01
C08-061416	L16060823-03	M7.062616.132202	06/26/16 13:22	01
C03-061416	L16060823-04	M7.062616.132434	06/26/16 13:24	01

Report Name: BLANK_SUMMARY
PDF File ID: 4828280
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060823 Prep Date: 06/16/16 07:27 Sample ID: WG572915-02
Instrument ID: CVAA1 Run Date: 06/26/16 12:36 Prep Method: 7470A
File ID: M7.062616.123627 Analyst: AMA Method: 7470A
Workgroup (AAB#): WG573236 Matrix: Water Units: mg/L
Contract #: Cal ID: CVAA1-26-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Mercury	0.000100	0.000400	0.000100	1	U

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name:BLANK
PDF ID: 4828281
28-JUN-2016 10:48

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060823 Run Date: 06/26/2016 Sample ID: WG572915-03
Instrument ID: CVAA1 Run Time: 12:38 Prep Method: 7470A
File ID: M7.062616.123858 Analyst: AMA Method: 7470A
Workgroup (AAB#): WG573236 Matrix: Water Units: mg/L
QC Key: DOD4 Lot#: STD76639 Cal ID: CVAA1-26-JUN-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Mercury	0.00400	0.00421	105	80 - 120	

LCS - Modified 03/06/2008
PDF File ID: 4828282
Report generated: 06/28/2016 10:48

Microbac Laboratories Inc.
MATRIX SPIKE AND MATRIX SPIKE DUP (MS/MSD)

Loginnum: L16060823 Cal ID: CVAA1 - Worknum: WG573236
Instrument ID: CVAA1 Contract #: Method: 7470A
Parent ID: WG572915-01 File ID: M7.062616.125410 Dil: 1 Matrix: WATER
Sample ID: WG572915-04 MS File ID: M7.062616.125643 Dil: 1 Units: mg/L
Sample ID: WG572915-05 MSD File ID: M7.062616.125137 Dil: 1

Analyte	Parent	MS Spiked	MS Found	MS %Rec	MSD Spiked	MSD Found	MSD %Rec	%RPD	%Rec Limits	RPD Limit	Q
Mercury	ND	0.00444	0.00479	108	0.00444	0.00480	108	0.0463	80 - 120	20	

* FAILS %REC LIMIT

\# FAILS RPD LIMIT

NOTE: This is an internal quality control sample.

WG_MS_MSD_DRYWT - Modified 05/26/2011
PDF File ID: 4828283
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
POST SPIKE REPORT

Sample Login ID: L16060823 **Worknum:** WG573236
Instrument ID: CVAA1 **Method:** 7470A
Post Spike ID: WG573236-01 **File ID:** M7.062616.124401 **Dil:** 1 **Units:** ug/L
Sample ID: L16060764-01 **File ID:** M7.062616.124129 **Dil:** 1 **Matrix:** Water

Analyte	Post Spike Result	C	Sample Result	C	Spike Added(SA)	% R	Control Limit %R	Q
MERCURY	1.07		0	U	1	107.1	85 - 115	

N = % Recovery exceeds control limits
F = Result is between MDL and RL
U = Sample result is below MDL. A value of zero is used in the calculation

POST_SPIKE - Modified 03/06/2008
PDF File ID: 4828278
Report generated: 06/28/2016 10:48

Microbac Laboratories Inc.

INITIAL CALIBRATION SUMMARY

Login Number: L16060823
Analytical Method: 7470A
ICAL Worknum: WG574322

Workgroup (AAB#): WG573236
Instrument ID: CVAA1
Initial Calibration Date: 06/26/2016 12:23

Analyte	WG574322-01		WG574322-02		WG574322-03		WG574322-04		WG574322-05		WG574322-06	
	STD	INT	STD	INT	STD	INT	STD	INT	STD	INT	STD	INT
Mercury	0	57.64	0.200	2673	1.00	13280	2.00	26640	5.00	66160	10.0	128600

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
* = Out of Compliance; R < 0.995

INT_CAL_HG_FU - Modified 03/06/2008
PDF File ID: 4828284
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.

INITIAL CALIBRATION SUMMARY

Login Number: L16060823
Analytical Method: 7470A
ICAL Worknum: WG574322

Workgroup (AAB#): WG573236
Instrument ID: CVAA1
Initial Calibration Date: 06/26/2016 12:23

Analyte	R	Q
Mercury	1.000	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
* = Out of Compliance; R < 0.995

INT_CAL_HG_FU - Modified 03/06/2008
PDF File ID: 4828284
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16060823
Instrument ID: CVAA1
File ID: M7.062616.122853
Workgroup (AAB#): WG573236
Matrix: WATER

Run Date: 06/26/2016
Run Time: 12:28
Analyst: AMA
Cal ID: CVAA1 - 26-JUN-16

Sample ID: WG574322-08
Method: 7470A
Units: ug/L

Analytes	MDL	RDL	Concentration	Qualifier
MERCURY	.1	.4	.1	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4828286
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/26/2016
Sample ID: WG574322-10
Instrument ID: CVAA1
Run Time: 12:33
Method: 7470A
File ID: M7.062616.123356
Analyst: AMA
Units: ug/L
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Mercury	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828288
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/26/2016
Sample ID: WG574322-12
Instrument ID: CVAA1
Run Time: 13:04
Method: 7470A
File ID: M7.062616.130420
Analyst: AMA
Units: ug/L
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Mercury	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828288
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060823
Run Date: 06/26/2016
Sample ID: WG574322-14
Instrument ID: CVAA1
Run Time: 13:34
Method: 7470A
File ID: M7.062616.133441
Analyst: AMA
Units: ug/L
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Mercury	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828288
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.

INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16060823 Run Date: 06/26/2016 Sample ID: WG574322-07
Instrument ID: CVAA1 Run Time: 12:26 Method: 7470A
File ID: M7.062616.122622 Analyst: AMA Units: ug/L
Workgroup (AAB#): WG573236 Cal ID: CVAA1 - 26-JUN-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Mercury		2	2.12	106	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4828285
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/26/2016
Sample ID: WG574322-09
Instrument ID: CVAA1
Run Time: 12:31
Method: 7470A
File ID: M7.062616.123125
Analyst: AMA
QC Key: DOD4
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Mercury, Total		0.00200	0.00210	mg/L	105	80 - 120		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828287
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/26/2016
Sample ID: WG574322-11
Instrument ID: CVAA1
Run Time: 13:01
Method: 7470A
File ID: M7.062616.130149
Analyst: AMA
QC Key: DOD4
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Mercury, Total		0.00200	0.00219	mg/L	109	80 - 120		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828287
Report generated 06/28/2016 10:48

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060823
Run Date: 06/26/2016
Sample ID: WG574322-13
Instrument ID: CVAA1
Run Time: 13:32
Method: 7470A
File ID: M7.062616.133211
Analyst: AMA
QC Key: DOD4
Workgroup (AAB#): WG573236
Cal ID: CVAA1 - 26-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Mercury, Total		0.00200	0.00223	mg/L	112	80 - 120		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828287
Report generated 06/28/2016 10:48

2.3.3.3 Raw Data

PDM CVAA1 245.1/7470/7471

Report Generated By CETAC QuickTrace
Analyst: VOA
Worksheet file: C:\Program Files (x86)\QuickTrace\Worksheets\062616A.wsz
Date Started: 6/26/2016 11:26:47 AM
Comment:

Results

Sample Name					Type	Date/Time	Conc (ug/L)	µAbs	%RSD	Flags	DF
Standard #0					STD	06/26/16 12:11:06 pm	0.0000	58	36.36		1.00
Replicates	72.2	26.9	61.9	69.6							
Standard #1 (0.2 ug/L)					STD	06/26/16 12:13:37 pm	0.2000	2673	0.35		1.00
Replicates	2686.9	2668.0	2667.5	2669.0							
Standard #2 (1.0 ug/L)					STD	06/26/16 12:16:09 pm	1.0000	13283	0.12		1.00
Replicates	13271.5	13294.2	13266.7	13300.2							
Standard #3 (2.0 ug/L)					STD	06/26/16 12:18:41 pm	2.0000	26641	0.01		1.00
Replicates	26642.8	26641.1	26641.1	26638.4							
Standard #4 (5.0 ug/L)					STD	06/26/16 12:21:14 pm	5.0000	66159	0.11		1.00
Replicates	66221.6	66218.8	66109.5	66085.3							
Standard #5 (10.0 ug/L)					STD	06/26/16 12:23:48 pm	10.0000	128636	0.07		1.00
Replicates	128716.2	128712.3	128583.1	128533.8							

Calibration

Equation: A = 499.952 + 12881.860C
R2: 0.99980
SEE: 796.4388
Flags:



ICV					ICV	06/26/16 12:26:22 pm	2.1170	27772	0.16		1.00
Replicates	27815.3	27806.0	27744.4	27723.2							
% Recovery	105.86										

Approved: June 28, 2016
Alexandra M. Alfred

Sample Name					Type	Date/Time	Conc (ug/L)	µAbs	%RSD	Flags	DF
ICB					ICB	06/26/16 12:28:53 pm	-0.0379	11	291.32		1.00
Replicates	4.1	-22.5	57.6	6.8							
CCV					CCV	06/26/16 12:31:25 pm	2.0950	27485	0.16		1.00
Replicates	27434.8	27461.4	27527.1	27517.9							
% Recovery	104.74										
CCB					CCB	06/26/16 12:33:56 pm	-0.0387	1	964.00		1.00
Replicates	9.0	5.0	-0.8	-9.8							
WG572915-02					UNK	06/26/16 12:36:27 pm	-0.0350	50	39.52		1.00
Replicates	72.8	57.2	40.4	27.9							
WG572915-03					UNK	06/26/16 12:38:58 pm	4.2070	54690	0.15		1.00
Replicates	54575.2	54727.8	54702.1	54755.0							
L1606076401					UNK	06/26/16 12:41:29 pm	-0.0379	11	72.00		1.00
Replicates	8.4	14.3	21.3	1.9							
WG573236-01					SPK	06/26/16 12:44:01 pm	1.0710	14290	0.11		1.00
Replicates	14266.8	14293.4	14303.5	14296.9							
% Recovery	110.84										
L1606076402					UNK	06/26/16 12:46:33 pm	-0.0345	56	29.18		1.00
Replicates	46.7	78.1	41.1	58.6							
L1606076403					UNK	06/26/16 12:49:05 pm	-0.0307	104	20.90		1.00
Replicates	76.8	97.6	116.7	125.9							
WG572915-01					UNK	06/26/16 12:51:37 pm	4.3160	56092	0.13		1.00
Replicates	56018.1	56175.3	56135.9	56040.3							
L1606076405					MSK	06/26/16 12:54:10 pm	-0.0402	-17	79.85	N	1.00
Replicates	-11.0	-38.2	-11.2	-9.2							
% Recovery	-108.89										
L1606076406					MSDUP	06/26/16 12:56:43 pm	4.3140	56069	0.23	D	1.00
Replicates	55903.2	56042.6	56133.8	56194.6							
% Recovery	-0.05	RPD 203.76									

Approved: June 28, 2016
Alexandra M. Alfred

Sample Name					Type	Date/Time	Conc (ug/L)	µAbs	%RSD	Flags	DF
L1606079401					UNK	06/26/16 12:59:16 pm	-0.0393	-7	153.33		1.00
Replicates	-17.5	2.6	1.3	-12.4							
CCV					CCV	06/26/16 01:01:49 pm	2.1870	28674	0.28		1.00
Replicates	28565.4	28662.0	28726.3	28743.6							
% Recovery	109.36										
CCB					CCB	06/26/16 01:04:20 pm	-0.0400	-15	71.00		1.00
Replicates	-20.6	-24.4	-16.6	0.3							
L1606079402					UNK	06/26/16 01:06:54 pm	-0.0360	36	40.24		1.00
Replicates	40.1	23.8	25.1	54.5							
L1606079403					UNK	06/26/16 01:09:27 pm	0.0214	775	1.64		1.00
Replicates	766.0	779.2	763.7	791.0							
L1606079404					UNK	06/26/16 01:11:58 pm	-0.0321	87	18.96		1.00
Replicates	68.6	84.0	86.7	108.7							
L1606082202					UNK	06/26/16 01:14:28 pm	-0.0340	62	33.50		1.00
Replicates	78.3	50.1	79.4	38.2							
L1606082203					UNK	06/26/16 01:16:59 pm	-0.0350	49	26.85		1.00
Replicates	46.5	64.4	52.2	32.7							
L1606082302					UNK	06/26/16 01:19:30 pm	-0.0247	181	8.96		1.00
Replicates	178.7	199.1	187.1	160.4							
L1606082303					UNK	06/26/16 01:22:02 pm	-0.0283	135	14.85		1.00
Replicates	148.4	150.3	135.4	106.8							
L1606082304					UNK	06/26/16 01:24:34 pm	-0.0349	51	26.20		1.00
Replicates	38.5	67.5	41.8	55.0							
L1606084001					UNK	06/26/16 01:27:06 pm	-0.0336	67	31.06		1.00
Replicates	78.3	84.6	37.9	65.6							

Approved: June 28, 2016
Alexandra M. Alfred

Sample Name					Type	Date/Time	Conc (ug/L)	µAbs	%RSD	Flags	DF
L1606084002					UNK	06/26/16 01:29:38 pm	-0.0355	42	37.05		1.00
Replicates	37.5	51.9	22.2	56.9							
CCV					CCV	06/26/16 01:32:11 pm	2.2310	29245	0.47		1.00
Replicates	29065.4	29210.7	29333.0	29372.9							
% Recovery	111.57										
CCB					CCB	06/26/16 01:34:41 pm	-0.0387	1	222.94		1.00
Replicates	7.9	10.6	18.4	-34.1							
L1606084003					UNK	06/26/16 01:37:14 pm	-0.0343	58	10.23		1.00
Replicates	58.3	56.5	65.5	51.2							
L1606084004					UNK	06/26/16 01:39:47 pm	-0.0311	99	14.17		1.00
Replicates	85.9	113.0	88.0	109.0							
L1606084005					UNK	06/26/16 01:42:20 pm	-0.0347	53	38.40		1.00
Replicates	51.7	58.0	26.5	75.7							
L1606084006					UNK	06/26/16 01:44:54 pm	-0.0362	34	52.68		1.00
Replicates	39.2	7.6	40.3	47.5							
CCV					CCV	06/26/16 01:47:26 pm	2.2480	29454	0.57		1.00
Replicates	29236.8	29421.4	29529.4	29626.5							
% Recovery	112.38										
CCB					CCB	06/26/16 01:49:57 pm	-0.0385	4	182.59		1.00
Replicates	14.9	-3.5	0.5	5.6							

Approved: June 28, 2016
Alexandra M. Alfred

3.0 Attachments

Microbac Laboratories Inc.
Ohio Valley Division Analyst List
July 8, 2016

001 - BIO-CHEM TESTING WVDEP 220
002 - REIC Consultants, Inc. WVDEP 060
003 - Sturm Environmental
004 - MICROBAC PITTSBURGH
005 - ES LABORATORIES
006 - ALCOSAN LABORATORIES
007 - ALS LABORATORIES
008 - BENCHMARK LABORATORIES
010 - MICROBAC CHICAGOLAND
AC - AMBER R. CARMICHAEL
ADC - ANTHONY D. CANTER
ADG - APRIL D. GREENE
AED - ALLEN E. DAVIS
ALS - ADRIANE L. STEED
AMA - ALEXANDRA M. ALFRED
AWE - ANDREW W. ESSIG
AZH - AFTER HOURS
BJO - BRIAN J. OGDEN
BKT - BRENDAN TORRENCE
BLG - BRENDA L. GREENWALT
BNB - Brandi N. Bentley
BRG - BRENDA R. GREGORY
CAA - CASSIE A. AUGENSTEIN
CAF - CHERYL A. FLOWERS
CEB - CHAD E. BARNES
CJQ - Cameron J. Quick
CJR - COURTNEY J. REXROAD
CLC - CHRYS L. CRAWFORD
CLS - CARA L. STRICKLER
CLW - CHARISSA L. WINTERS
CPD - CHAD P. DAVIS
CSH - CHRIS S. HILL
CV - Carl Volkman
DAK - DEAN A. KETELSEN
DCM - DAVID C. MERCKLE
DEV - DAVID E. VANDENBERG
DIH - DEANNA I. HESSON
DLB - DAVID L. BUMGARNER
DLP - DOROTHY L. PAYNE
DLW - DIANA L. WRIGHT
DSM - DAVID S. MOSSOR
ECL - ERIC C. LAWSON
EMW - ERIC M. WILKEN
ENY - EMILY N. YOAK
ERP - ERIN R. PORTER
FJB - FRANCES J. BOLDEN
JBK - JEREMY B. KINNEY
JDH - JUSTIN D. HESSON
JDS - JARED D. SMITH
JJS - JOHN J. STE MARIE
JKP - JACQUELINE K. PARSONS
JLD - JESSICA L. DELONG
JLL - JOHN L. LENT
JMW - JEANA M. WHITE
JTP - JOSHUA T. PEMBERTON
JWR - JOHN W. RICHARDS
JWS - JACK W. SHEAVES
JYH - JI Y. HU
KAJ - KELLIE A. JOHNSON
KAT - KATHY A. TUCKER
KDW - KATHRYN D. WELCH
KEB - KATIE E. BARNES
KHR - KIM H. RHODES
KKB - KERRI K. BUCK
KRA - KATHY R. ALBERTSON
KRB - KAELY R. BECKER
KRP - KATHY R. PARSONS
LEC - LAURA E. CARPENTER
LJH - Lacey J. Hendershot
LKN - LINDA K. NEDEFF
LLS - LARRY L. STEPHENS
LSB - LESLIE S. BUCINA
MAP - MARLA A. PORTER
MBK - MORGAN B. KNOWLTON
MDA - MIKE D. ALBERTSON
MDC - MIKE D. COCHRAN
MES - MARY E. SCHILLING
MLB - MEGAN L. BACHE
MMB - MAREN M. BEERY
MRT - MICHELLE R. TAYLOR
MSW - MATT S. WILSON
NPH - Natalie P. Hart
PDM - PIERCE D. MORRIS
PIT - MICROBAC WARRENDALE
PRL - PAIGE R. LAMB
PSW - PEGGY S. WEBB
QX - QIN XU
RAH - ROY A. HALSTEAD
REK - BOB E. KYER
RLB - BOB BUCHANAN
RM - RAYMOND MALEKE
RNP - RICK N. PETTY
RST - ROBIN S. TURNER
SAV - SARAH A. VANDENBERG
SCB - SARAH C. BOGOLIN
SDC - SHALYN D. CONLEY
SLM - STEPHANIE L. MOSSBURG
SLP - SHERI L. PFALZGRAF
TB - TODD BOYLE
TGF - TIM G. FELTON
TMB - TIFFANY M. BAILEY
TMM - TAMMY M. MORRIS
VC - VICKI COLLIER
WJB - WILL J. BEASLEY
WRR - WESLEY R. RICHARDS
WTD - WADE T. DELONG
XXX - UNAVAILABLE OR SUBCONTRACT

Microbac Laboratories Inc.
List of Valid Qualifiers
July 08, 2016

Qualkey: DOD

Qualifier	Description
*	Surrogate or spike compound out of range
+	Correlation coefficient for the MSA is less than 0.995
<	Result is less than the associated numerical value.
>	Greater than
A	See the report narrative
B	The reported result is associated with a contaminated method blank.
B,H1	Analyte present in method blank. Sample analysis performed past holding time.
B1	Target analyte detected in method blank at or above the method reporting limit
B3	Target analyte detected in calibration blank at or above the method reporting limit
B4	The BOD unseeded dilution water blank exceeded 0.2 mg/L
C	Confirmed by GC/MS
CG	Confluent growth
CT1	Cooler temperature at sample reciept exceeded regulatory limit.
DL	Surrogate or spike compound was diluted out
E	Estimated concentration due to sample matrix interference
E,CT1	Estimated results. The cooler temperature at receipt exceeded regulatory guidelines for requested testing.
EDL	Elevated sample reporting limits, presence of non-target analytes
EMPC	Estimated Maximum Possible Concentration
F, S	Estimated result below quantitation limit; method of standard additions(MSA)
F,CT1	Estimated value; the analyte concentration was less than the RL/LOQ. The cooler temperature at receipt exceeded regula
FL	Free Liquid
FP1	Did not ignite.
H1	Sample analysis performed past holding time.
H1,CT1	Sample analysis performed past holding time. The cooler temperature at receipt exceeded regulatory guidelines for reque
I	Semiquantitative result (out of instrument calibration range)
J	Estimated concentration; sample matrix interference.
J	Estimated value ; the analyte concentration was greater than the highest standard
J	Estimated value ; the analyte concentration was less than the LOQ.
J	The reported result is an estimated value.
J,B	Analyte detected in both the method blank and sample above the MDL.
J,CT1	Estimated value ; the analyte concentration was less than the LOQ. Cooler temperature at sample reciept exceeded regu
J,H1	Estimated value ; the analyte concentration was less than the LOQ. Sample analysis performed past holding time.
J,H1	The reported result is an estimated value. Sample was analyzed past holding time.
J,P	Estimate; columns don't agree to within 40%
J,S	Estimated concentration; analyzed by method of standard addition (MSA)
JB	The reported result is an estimated value. The reported result is also associated with a contaminated method blank.
JQ	The reported result is an estimated value and one or more quality control criteria failed. See narrative.
L	Sample reporting limits elevated due to matrix interference
L1	The associated blank spike (LCS) recovery was above the laboratory acceptance limits.
L2	The associated blank spike (LCS) recovery was below the laboratory acceptance limits.
M	Matrix effect; the concentration is an estimate due to matrix effect.
N	Nontarget analyte; the analyte is a tentativlely identified compound (TIC) by GC/MS
NA	Not applicable
ND	Not detected at or above the reporting limit (RL)
ND, B	Not detected at or above the reporting limit (RL). Analyte present in method blank.
ND, CT1	Analyte was not detected. The concentration is below the reported LOD. The cooler temperature at receipt exceeded reg
ND, L	Not detected; sample reporting limit (RL) elevated due to interference
ND, S	Not detected; analyzed by method of standard addition (MSA)
ND,H1	Not detected; Sample analysis performed past holding time.
ND,H1,CT1	Not detected; Sample analysis performed past holding time. The cooler temperature at receipt exceeded regulatory guide
NF	Not found by library search
NFL	No free liquid
NI	Non-ignitable
NR	Analyte is not required to be analyzed
NS	Not spiked
P	Concentrations >40% difference between the two GC columns
Q	One or more quality control criteria failed. See narrative.
Q,H1	One or more quality control criteria failed. Sample analyzed past holding time. See narrative.
QNS	Quantity of sample not sufficient to perform analysis
RA	Reanalysis confirms reported results
RE	Reanalysis confirms sample matrix interference
S	Analyzed by method of standard addition (MSA)
SMI	Sample matrix interference on surrogate
SP	Reported results are for spike compounds only
T5	Laboratory not licensed for this parameter
TIC	Library Search Compound
TNTC	Too numerous to count

Microbac Laboratories Inc.
List of Valid Qualifiers
July 08, 2016

Qualkey: DOD

TNTC, B	Too numerous to count. Analyte present in method blank.
TNTC,CT1	Too numerous to count. The cooler temperature at receipt exceeded regulatory guidelines for requested testing.
TNTC,H1	Too numerous to count. Sample analysis performed past holding time.
U	Analyte was not detected. The concentration is below the reported LOD.
U,CT1	Analyte was not detected. The concentration is below the reported LOD. Cooler temperature at sample reciept exceeded
U,H1	Not detected; Sample analysis performed past holding time.
UJ	Undetected; the MDL and RL are estimated due to quality control discrepancies.
UQ	Undetected; the analyte was analyzed for, but not detected.
W	Post-digestion spike for furnace AA out of control limits
X	Exceeds regulatory limit
X, S	Exceeds regulatory limit; method of standard additions (MSA)
Z	Cannot be resolved from isomer - see below

AECOM

Chain of Custody Record

COC Number:

Laboratory:	Microbac POC: Stephanie Mossburg	Project Manager:	Debra Richmann
Address:	158 Starlite Drive	Phone/Fax Number:	210-296-2000
	Marietta, OH 45750	Sampler (print):	Scott Beesinger
Phone:	1-800-373-4071		
Client:	AECOM	Signature:	Scott Beesinger
Address:	112 East Pecan Ste. 400		
	San Antonio, TX 78205		
Turn Around Time:	STANDARD	pH:	
Project Name/Location:	Longhorn		
Project Number:	60274185.0013BH 1824 MR10		

Mail to: Linda Raabe, 112 East Pecan STE. 400, San Antonio, TX 78205, 210-296-2000

Fed Ex Airbill No:

Program:

ERPIMS REQUIRED FIELDS

Site Name	Sample ID/Location ID	SBD	SED	Date	Time	Comp*	Grab	Matrix	Number of Containers	VOC	Perchlorate	Total Metals	1,4 - DIOXANE					SA CODE	Cooler ID	LOT CONTROL NUMBERS ABLOT	EBLOT	TBLOT
SITE 18/24	126-061416			6/14/16	1110		✓	W	6	✓	✓		✓									
	126F-061416			6/14/16	1110		✓	W	1			✓										
	C08-061416			6/14/16	1310		✓	W	7	✓	✓	✓	✓									
	C03-061416			6/14/16	1415		✓	W	7	✓	✓	✓	✓									
	TRIP BLANK			6/14/16			✓	W	2	✓												

Comments: **STANDARD TAT**

	Date	Time	Received by: (Signature)	Received by: (Signature)	Date	Time	Relinquished by: (Signature)
Relinquished by: (Signature) Scott Beesinger	6/14/16	1530					
Relinquished by: (Signature)	Date	Time	Rece (Sign			Time	Remarks:

Microbac OVD
Received: 06/15/2016 09:56
By: BRENDA GREGORY
221000087199

Brenda Gregory

*Homogenize all composite samples prior to analysis

ager, Pink QA/QC Manager

00714138

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060823
Account: 2551
Project: 2551.096
Samples: 5
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060823-01	760859	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:15	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:15	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:15	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Samplenum	Container ID	Products
L16060823-01	760860	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	PREP	W1	EXT	15-JUN-2016 17:32	CPD	AZH	
3	DISP	EXT	DISP	16-JUN-2016 15:22	BJO	BJO	
4	ANALYZ*	EXT	SEMI	20-JUN-2016 09:18	NPH	CPD	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	STORE	W1	A1	24-JUN-2016 08:23	BRG	BRG	

Sample extract/digestate/leachate

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060823
Account: 2551
Project: 2551.096
Samples: 5
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060823-01	760861	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	ANALYZ	W1	SEM	24-JUN-2016 12:12	JWR	BRG	
3	STORE	SEM	A1	29-JUN-2016 13:20	CLS	JWR	

Samplenum	Container ID	Products
L16060823-02	760862	AG-MS AL AS-MS BA-MS BE-AX CA CD-MS CO-MS CR-M

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	PREP	W1	DIG	15-JUN-2016 14:59	AC	CLS	
3	ANALYZ*	DIG	METALS	17-JUN-2016 14:30	CJQ	AC	
4	STORE	DIG	A1	17-JUN-2016 14:50	CLS	ERP	

Sample extract/digestate/leachate

Samplenum	Container ID	Products
L16060823-03	760863	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:15	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:16	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060823
Account: 2551
Project: 2551.096
Samples: 5
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060823-03	760864	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	PREP	W1	EXT	15-JUN-2016 17:32	CPD	AZH	
3	DISP	EXT	DISP	16-JUN-2016 15:22	BJO	BJO	
4	ANALYZ*	EXT	SEMI	20-JUN-2016 09:18	NPH	CPD	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	STORE	W1	A1	24-JUN-2016 08:23	BRG	BRG	

Sample extract/digestate/leachate

Samplenum	Container ID	Products
L16060823-03	760865	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	ANALYZ	W1	SEM	24-JUN-2016 12:12	JWR	BRG	
3	STORE	SEM	A1	29-JUN-2016 13:20	CLS	JWR	

Samplenum	Container ID	Products
L16060823-03	760866	HG K MG MN-MS NA NI-MS PB-MS SB-MS SE-AX TL-MS

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	PREP	W1	DIG	15-JUN-2016 14:59	AC	CLS	
3	ANALYZ*	DIG	METALS	17-JUN-2016 14:30	CJQ	AC	
4	STORE	DIG	A1	17-JUN-2016 14:50	CLS	ERP	

Sample extract/digestate/leachate

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060823
Account: 2551
Project: 2551.096
Samples: 5
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060823-04	760867	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:15	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:15	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:15	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Samplenum	Container ID	Products
L16060823-04	760868	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	PREP	W1	EXT	15-JUN-2016 17:32	CPD	AZH	
3	DISP	EXT	DISP	16-JUN-2016 15:23	BJO	BJO	
4	ANALYZ*	EXT	SEMI	20-JUN-2016 09:18	NPH	CPD	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	STORE	W1	A1	24-JUN-2016 08:23	BRG	BRG	

Sample extract/digestate/leachate

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060823
Account: 2551
Project: 2551.096
Samples: 5
Due Date: 24-JUN-2016

Samplenum	Container ID	Products
L16060823-04	760869	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	ANALYZ	W1	SEM	24-JUN-2016 12:12	JWR	BRG	
3	STORE	SEM	A1	29-JUN-2016 13:20	CLS	JWR	

Samplenum	Container ID	Products
L16060823-04	760870	AG-MS AL AS-MS BA-MS BE-AX CA CD-MS CO-MS CR-M

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	15-JUN-2016 14:53	BRG		
2	PREP	W1	DIG	15-JUN-2016 14:59	AC	CLS	
3	ANALYZ*	DIG	METALS	17-JUN-2016 14:30	CJQ	AC	
4	STORE	DIG	A1	17-JUN-2016 14:50	CLS	ERP	

Sample extract/digestate/leachate

Samplenum	Container ID	Products
L16060823-05	760871	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:15	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	15-JUN-2016 14:53	BRG		
2	ANALYZ	V1	ORG4	16-JUN-2016 08:15	AWE	CLS	
3	STORE	ORG4	A1	29-JUN-2016 12:15	CLS	JST	

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

NELAP Addendum - January 4, 2016

Non-NELAP LIMS Product and Description

The following is a list of those tests that are not included in the Microbac – OVD NELAP Scope of Accreditation:

Heat of Combustion (BTU)
Total Halide by Bomb Combustion (TX)
Particle Sizing - 200 Mesh (PS200)
Specific Gravity/Density (SPGRAV)
Total Residual Chlorine (CL-TRL)
Total Volatile Solids (all forms) (TVS)
Total Coliform Bacteria (all methods)
Fecal Coliform Bacteria (all methods)
Sulfite (SO3)
Propionaldehyde (HPLC-UV)

SOLID AND HAZARDOUS CHEMICALS
Nitrogen, Ammonia by Method 350.1
Chromium, Hexavalent, Leachable by SM3500 Cr-B 2009
Phenolics, Total by Method 420.1
ASTM D3987-06

NELAP Accreditation by Laboratory SOP

NONPOTABLE WATER

OVD HPLC02/HPLC-UV

Nitroglycerin
Acetic acid
Butyric acid
Lactic acid
Propionic acid
Pyruvic acid

OVD MSS01/GC-MS

1,4-Phenylenediamine
1-Methylnaphthalene
1,4-Dioxane
Atrazine
Benzaldehyde
Biphenyl
Caprolactam
Hexamethylphosphoramide (HMPA)
Pentachlorobenzene
Pentachloroethane

NELAP Accreditation by Laboratory SOP

NONPOTABLE WATER

OVD MSV01/GC-MS

1, 1, 2-Trichloro-1,2,2-trifluoroethane
1,3-Butadiene
Cyclohexane
Cyclohexanone
Dimethyl disulfide
Dimethylsulfide
Ethyl-t-butylether (ETBE)
Isoprene
Methylacetate
Methylcyclohexane
T-amylmethylether (TAME)
Tetrahydrofuran (THF)

OVD HPLC07/HPLC-MS-MS

Hexamethylphosphoramide (XMPA-LCMS)

OVD HPLC12/HPLC/UV

Acetate
Formate

OVD RSK01/GC-FID

Acetylene
Propane

OVD K9305/ISE

Fluoroborate

SOLID AND HAZARDOUS CHEMICALS

OVD MSS0I/GC-MS

1-Methylnaphthalene
Benzaldehyde
Biphenyl
Caprolactam
Pentachloroethane

NELAP Accreditation by Laboratory SOP

SOLID AND HAZARDOUS CHEMICALS

OVD MSV0I/GC-MS

1.3-Butadiene
Cyclohexane
Cyclohexanone
Dimethyl disulfide
Dimethylsulfide
Ethyl-t-butylether (ETBE)
Isoprene
Methylacetate
Methylcyclohexane
n-Hexane
T-amylmethylether (TAME)

Microbac

Laboratory Report Number: L16060927

Kayla Teague
AECOM Technical Services, Inc.
16000 Dallas Parkway
Dallas, TX 75248

Please find enclosed the analytical results for the samples you submitted to Microbac Laboratories. Review and compilation of your report was completed by Microbac's Ohio Valley Division (OVD). If you have any questions, comments, or require further assistance regarding this report, please contact your service representative listed below.

Laboratory Contact:
Michelle Taylor – Client Services Specialist
(740) 373-4071
Michelle.Taylor@microbac.com

I certify that all test results meet all of the requirements of the DoD QSM and other applicable contract terms and conditions. Any exceptions are attached to this cover page or addressed in the method narratives presented in the report. All results for soil samples are reported on a 'dry-weight' basis unless specified otherwise. Analytical results for water and wastes are reported on a 'as received' basis unless specified otherwise. A statement of uncertainty for each analysis is available upon request. This laboratory report shall not be reproduced, except in full, without the written approval of Microbac Laboratories, DoD ELAP certification number 2936.01. The reported results are related only to the samples analyzed as received.

This report was certified on July 06 2016

David Vandenberg – Managing Director

State of Origin: TX
Accrediting Authority: Texas Commission on Environmental Quality ID:T104704252-07-TX
QAPP: DOD Ver 4.1

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Record of Sample Receipt and Inspection

Comments/Discrepancies

This is the record of the shipment conditions and the inspection records for the samples received and reported as a sample delivery group (SDG). All of the samples were inspected and observed to conform to our receipt policies, except as noted below.

There were no discrepancies.

Discrepancy	Resolution

Coolers

Cooler #	Temperature Gun	Temperature	COC #	Airbill #	Temp Required?
0015011	H	6.0		J2317164993	X

Inspection Checklist

#	Question	Result
1	Were shipping coolers sealed?	Yes
2	Were custody seals intact?	Yes
3	Were cooler temperatures in range of 0-6?	Yes
4	Was ice present?	Yes
5	Were COC's received/information complete/signed and dated?	Yes
6	Were sample containers intact and match COC?	Yes
7	Were sample labels intact and match COC?	Yes
8	Were the correct containers and volumes received?	Yes
9	Were samples received within EPA hold times?	Yes
10	All samples were checked for pH and met the standard. Exceptions are noted above under discrepancy. (water only)	Yes
11	Were pH ranges acceptable? (voa's excluded)	NA
12	Were VOA samples free of headspace (less than 6mm)?	Yes

Microbac Laboratories ● Ohio Valley Division
158 Starlite Drive, Marietta, OH 45750 ● T: (740)373-4071 F: (740)373-4835
www.microbac.com

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Samples Received

Client ID	Laboratory ID	Date Collected	Date Received
18CPTMW23SW-061516	L16060927-01	06/15/2016 08:00	06/16/2016 10:10
18CPTMW23-061516	L16060927-02	06/15/2016 09:05	06/16/2016 10:10
18CPTMW19SW-061516	L16060927-03	06/15/2016 10:15	06/16/2016 10:10
18CPTMW19-061516	L16060927-04	06/15/2016 11:15	06/16/2016 10:10
TRIP BLANK	L16060927-05	06/15/2016 00:01	06/16/2016 10:10

Microbac REPORT L16060927
PREPARED FOR AECOM Technical Services, Inc.
WORK ID:

1.0 Summary Data 5
1.1 Narratives 6
1.2 Certificate of Analysis 25
2.0 Full Sample Data Package 49
2.1 Volatiles Data 50
2.1.1 Volatiles GCMS Data (8260) 51
2.1.1.1 Summary Data 52
2.1.1.2 QC Summary Data 69
2.1.1.3 Sample Data 125
2.1.1.4 Standards Data 163
2.1.1.5 Raw QC Data 230
2.2 Semivolatiles Data 356
2.2.1 Semivolatiles GC/MS Data (Dioxane) 260
2.2.1.1 Summary Data 261
2.2.1.2 QC Summary Data 268
2.2.1.3 Sample Data 298
2.2.1.4 Standards Data 313
2.2.1.5 Raw QC Data 337
2.2 General Chromatography Data 356
2.2.2 6850 LC/MS Data 357
2.2.2.1 Summary Data 358
2.2.2.2 QC Summary Data 361
2.2.2.3 Sample Data 432
2.2.2.4 Standards Data 441
2.2.2.5 Raw QC Data 502
3.0 Attachments 509

1.0 Summary Data

1.1 Narratives

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	8260
Prep Batch Number(s):	573645, 573771, 573991	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-06 00:00:00		

Laboratory Data Package Cover Page

X	R1	Field chain-of-custody documentation;
X	R2	Sample identification cross-reference;
X	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
X	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
X	R5	Test reports/summary forms for blank samples;
X	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
X	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
X	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
X	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
X	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Franci Bolden	Franci Bolden	Analyst I	2016-07-06 17:10:40

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	8260
Prep Batch Number(s):	573645, 573771, 573991	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-06 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports					
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?	X				
Were % moisture (or solids) reported for all soil and sediment samples?	X				
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?	X				
Were surrogate percent recoveries in all samples within the laboratory QC limits?	X				
Test reports/summary forms for blank samples	X				
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):					
Were all COCs included in the LCS?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	8260
Prep Batch Number(s):	573645, 573771, 573991	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-06 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?	X				
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?	X				
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	8260
Prep Batch Number(s):	573645, 573771, 573991	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-06 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?	X				
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?		X			1
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?			X		
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?	X				
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?	X				
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?			X		
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?			X		
Method detection limit (MDL) studies					

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	8260
Prep Batch Number(s):	573645, 573771, 573991	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-06 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	8260
Prep Batch Number(s):	573645, 573771, 573991	**Reviewer Name:**	Franci Bolden
LRC Date:	2016-07-06 00:00:00		

below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

1) Bromomethane was below the LCL and Carbon Tetrachloride, 2,2-Dichloropropane and Trichlorofluoromethane exceeded the UCL in CCV WG573770-02 analyzed 06/22/2016 on HPMS8.
2,2-Dichloropropane exceeded the UCL in CCV WG573990-02 analyzed 06/23/2016 on HPMS8.

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-28 00:00:00		

Laboratory Data Package Cover Page

X	R1	Field chain-of-custody documentation;
X	R2	Sample identification cross-reference;
X	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
X	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
X	R5	Test reports/summary forms for blank samples;
X	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
X	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
X	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
X	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
X	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Eric Lawson		Chemist III	2016-06-28 13:53:35

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-28 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports					
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?			X		
Were % moisture (or solids) reported for all soil and sediment samples?			X		
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?	X				
Were surrogate percent recoveries in all samples within the laboratory QC limits?		X			1
Test reports/summary forms for blank samples	X				
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):					
Were all COCs included in the LCS?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-28 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?	X				
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?	X				
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-28 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?	X				
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?	X				
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?			X		
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?	X				
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?	X				
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?			X		
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?			X		
Method detection limit (MDL) studies					

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-28 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-28 00:00:00		

below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

1. Sample 02 yielded a recovery for the surrogate that was above the acceptance criteria. This sample was run at a dilution and the surrogate result was within acceptance limits, indicating matrix interference.

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Laboratory Data Package Cover Page

	R1	Field chain-of-custody documentation;
	R2	Sample identification cross-reference;
	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
	R5	Test reports/summary forms for blank samples;
	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Wade DeLong	Wade DeLong	Chemist I	2016-06-28 17:46:14

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports					
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?			X		
Were % moisture (or solids) reported for all soil and sediment samples?			X		
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?			X		
Were surrogate percent recoveries in all samples within the laboratory QC limits?			X		
Test reports/summary forms for blank samples	X				
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):					
Were all COCs included in the LCS?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?	X				
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?			X		
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?	X				
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?	X				
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?	X				
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?			X		
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?	X				
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?	X				
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?			X		
Method detection limit (MDL) studies					

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16060927
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

1.2 Certificate of Analysis

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-01	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	18CPTMW23SW-061516	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573991	**Analyst:**	FJB	**Run Date:**	06/24/2016 00:24
Collect Date:	06/15/2016 08:00	**Dilution:**	1	**File ID:**	8M413199
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	9.67	J	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	U	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	111	85	115	
1,2-Dichloroethane-d4	116	70	120	
Toluene-d8	104	85	120	
4-Bromofluorobenzene	101	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-01	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	18CPTMW23SW-061516	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 15:30
Collect Date:	06/15/2016 08:00	**Dilution:**	1	**File ID:**	15M18514
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	32.4	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-01	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	18CPTMW23SW-061516	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 19:02
Collect Date:	06/15/2016 08:00	**Dilution:**	1	**File ID:**	1LM.LM35501
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	0.407		0.400	0.200	0.100

Certificate of Analysis

Sample #:	L16060927-02	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	18CPTMW23-061516	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573771	**Analyst:**	FJB	**Run Date:**	06/22/2016 18:40
Collect Date:	06/15/2016 09:05	**Dilution:**	1	**File ID:**	8M413169
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	1.59		1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	Q	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	5.45		1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	1.06		1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	214		1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	10.1		2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	232		1.00	0.500	0.250
trans-1,2-Dichloroethene	156-60-5	4.09		1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.256	J	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	1.22		1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.858	J	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	Q	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	5.89		1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	112	85	115	
1,2-Dichloroethane-d4	115	70	120	
Toluene-d8	105	85	120	
4-Bromofluorobenzene	99.4	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-02	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	18CPTMW23-061516	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573991	**Analyst:**	FJB	**Run Date:**	06/23/2016 22:56
Collect Date:	06/15/2016 09:05	**Dilution:**	200	**File ID:**	8M413196
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Trichloroethene	79-01-6	3820		200	100	50.0

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	112	85	115	
1,2-Dichloroethane-d4	116	70	120	
Toluene-d8	101	85	120	
4-Bromofluorobenzene	101	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-02	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	18CPTMW23-061516	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 15:53
Collect Date:	06/15/2016 09:05	**Dilution:**	1	**File ID:**	15M18515
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	7.62	J	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	133	20	129	*

J	The reported result is an estimated value.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-02	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	18CPTMW23-061516	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/25/2016 12:46
Collect Date:	06/15/2016 09:05	**Dilution:**	10	**File ID:**	15M18553
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	8.02	J	20.0	10.0	5.00

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	109	20	129	

J	Estimated value ; the analyte concentration was less than the LOQ.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-02	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	18CPTMW23-061516	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 19:21
Collect Date:	06/15/2016 09:05	**Dilution:**	1000	**File ID:**	1LM.LM35502
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	2310		400	200	100

Certificate of Analysis

Sample #:	L16060927-03	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	18CPTMW19SW-061516	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573991	**Analyst:**	FJB	**Run Date:**	06/24/2016 00:53
Collect Date:	06/15/2016 10:15	**Dilution:**	1	**File ID:**	8M413200
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.388	J	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	U	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	0.599	J	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	J	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.506	J	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.484	J	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	113	85	115	
1,2-Dichloroethane-d4	119	70	120	
Toluene-d8	104	85	120	
4-Bromofluorobenzene	103	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-03	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	18CPTMW19SW-061516	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 16:15
Collect Date:	06/15/2016 10:15	**Dilution:**	1	**File ID:**	15M18516
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	37.5	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-03	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	18CPTMW19SW-061516	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/24/2016 23:18
Collect Date:	06/15/2016 10:15	**Dilution:**	1	**File ID:**	1LM.LM35473
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	1.25		0.400	0.200	0.100

Certificate of Analysis

Sample #:	L16060927-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	18CPTMW19-061516	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573645	**Analyst:**	FJB	**Run Date:**	06/22/2016 03:34
Collect Date:	06/15/2016 11:15	**Dilution:**	1	**File ID:**	8M413145
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	U	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	112	85	115	
1,2-Dichloroethane-d4	114	70	120	
Toluene-d8	104	85	120	
4-Bromofluorobenzene	99.9	75	120	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	18CPTMW19-061516	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 16:38
Collect Date:	06/15/2016 11:15	**Dilution:**	1	**File ID:**	15M18517
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	40.5	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-04	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	18CPTMW19-061516	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/24/2016 23:37
Collect Date:	06/15/2016 11:15	**Dilution:**	1	**File ID:**	1LM.LM35474
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	5.31		0.400	0.200	0.100

Certificate of Analysis

Sample #:	L16060927-05	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	TRIP BLANK	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573645	**Analyst:**	FJB	**Run Date:**	06/21/2016 22:44
Collect Date:	06/15/2016 00:01	**Dilution:**	1	**File ID:**	8M413135
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	U	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	111	85	115	
1,2-Dichloroethane-d4	112	70	120	
Toluene-d8	103	85	120	
4-Bromofluorobenzene	100	75	120	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

2.0 Full Sample Data Package

2.1 Volatiles Data

2.1.1 Volatiles GCMS Data (8260)

2.1.1.1 Summary Data

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-01	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	18CPTMW23SW-061516	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573991	**Analyst:**	FJB	**Run Date:**	06/24/2016 00:24
Collect Date:	06/15/2016 08:00	**Dilution:**	1	**File ID:**	8M413199
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	9.67	J	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	U	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	111	85	115	
1,2-Dichloroethane-d4	116	70	120	
Toluene-d8	104	85	120	
4-Bromofluorobenzene	101	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-02	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	18CPTMW23-061516	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573771	**Analyst:**	FJB	**Run Date:**	06/22/2016 18:40
Collect Date:	06/15/2016 09:05	**Dilution:**	1	**File ID:**	8M413169
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	1.59		1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	Q	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	5.45		1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	1.06		1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	214		1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	10.1		2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	232		1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	4.09		1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.256	J	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	1.22		1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.858	J	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	Q	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	5.89		1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	112	85	115	
1,2-Dichloroethane-d4	115	70	120	
Toluene-d8	105	85	120	
4-Bromofluorobenzene	99.4	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-02	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	18CPTMW23-061516	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573991	**Analyst:**	FJB	**Run Date:**	06/23/2016 22:56
Collect Date:	06/15/2016 09:05	**Dilution:**	200	**File ID:**	8M413196
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Trichloroethene	79-01-6	3820		200	100	50.0

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	112	85	115	
1,2-Dichloroethane-d4	116	70	120	
Toluene-d8	101	85	120	
4-Bromofluorobenzene	101	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-03	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	18CPTMW19SW-061516	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573991	**Analyst:**	FJB	**Run Date:**	06/24/2016 00:53
Collect Date:	06/15/2016 10:15	**Dilution:**	1	**File ID:**	8M413200
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.388	J	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	U	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	0.599	J	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	J	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.506	J	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	Q	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.484	J	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	113	85	115	
1,2-Dichloroethane-d4	119	70	120	
Toluene-d8	104	85	120	
4-Bromofluorobenzene	103	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	18CPTMW19-061516	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573645	**Analyst:**	FJB	**Run Date:**	06/22/2016 03:34
Collect Date:	06/15/2016 11:15	**Dilution:**	1	**File ID:**	8M413145
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	U	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	112	85	115	
1,2-Dichloroethane-d4	114	70	120	
Toluene-d8	104	85	120	
4-Bromofluorobenzene	99.9	75	120	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-05	**PrePrep Method:**	N/A	**Instrument:**	HPMS8
Client ID:	TRIP BLANK	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 17:50
Workgroup #:	WG573645	**Analyst:**	FJB	**Run Date:**	06/21/2016 22:44
Collect Date:	06/15/2016 00:01	**Dilution:**	1	**File ID:**	8M413135
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	U	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	U	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	111	85	115	
1,2-Dichloroethane-d4	112	70	120	
Toluene-d8	103	85	120	
4-Bromofluorobenzene	100	75	120	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

2.1.1.2 QC Summary Data

Example 8260 Calculations

1.0 Calculating the Response Factor (RF) from the initial calibration (ICAL) data:

RF = [(Ax) (Cis)] / [(Ais) (Cx)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured:	3399156
Cis = Concentration of the specific internal standard (ug/mL)	25
Ais = Area of the characteristic ion of the specific internal standard	846471
Cx = Concentration of the compound in the standard being measured (ug/mL)	100
RF = Calculated Response Factor	**1.0039**

2.0 Calculating the concentration (C) of a compound in water using the average RF: *

Cx = [(Ax) (Cis) (Vn)(D)] / [(Ais) (RF) (Vs)]

where:	**Example**
Ax = Area of the characteristic ion for the compound being measured	**3122498**
Cis = Concentration of the specific internal standard (ug/L)	**25**
D = Dilution factor for sample as a multiplier (10x = 10)	**1**
Ais = Area of the characteristic ion of the specific internal standard	**611048**
RF = Average RF from the ICAL	**1.004**
Vs = Purge volume of sample (mL)	**10**
Vn = Nominal purge volume of sample (mL) (10.0 mL)	**10**
Cx = Concentration of the compound in the sample being measured (ug/L)	**127.2428**

3.0 Calculating the concentration (C) of a compound in soil using the average RF: *

Cx = [(Ax) (Cis) (Wn)(D)] / [(Ais) (RF) (Ws)]

where:	**Example**
Ax = Area of the characteristic ion for the compound being measured	**3122498**
Cis = Concentration of the specific internal standard (ug/L)	**25**
D = Dilution factor for sample as a multiplier (10x = 10)	**1**
Ais = Area of the characteristic ion of the specific internal standard	**611048**
RF = Average RF from the ICAL	**1.004**
Ws = Weight of sample purged (g)	**5**
Wn = Nominal purge weight (g) (5.0 g)	**5**
Cx = Concentration of the compound in the sample being measured (ug/L)	**127.2428**

Dry weight correction:

Percent solids (PCT_S)	**50**
Cd = (Cx) (100)/PCT_S	**254.4856**

* Concentrations appearing on the instrument quantitation reports are on-column results and do not take into account initial volume, final volume, and the dilution factor.

4.0 Concentration from Linear Regression

Step 1: Retrieve Curve Data From Plot, y = mx + b

y = response ratio = response of analyte / response of IS = Ax/Ais

x = amount ratio = concentration analyte/concentration internal standard = Cx / Cis

m = slope from curve = 0.213

b = intercept from curve = - 0.00642

Step 2: Calculate y from Quantitation Report

y = 86550/593147 = 0.1459

Step 3: Solve for x

x = (y - b)/m = [(0.1459 -(- 0.00642)]/ 0.213 = 0.7152

Step 4: Solve for analyte concentration Cx

Cx = Cis (x) = (25.0)(0.7152) = 17.88

Example Spreadsheet Calculation:

Slope from curve, m:	**0.213**
Intercept from curve, b:	**-0.00642**
Area of analyte, Ax:	**86550**
Area of Internal Standard , Ais:	**593147**
Concentration of IS, Cis	**25.00**
Response Ratio:	**0.145917**
Amount Ratio:	**0.715195**
Concentration:	**17.87988**
Units of Internal Standard:	**ug/L**

5.0 Concentration from Quadratic Regression

Step 1 - Retrieve Curve Data from Plot, y = Ax^2 + Bx + C
Where:
Ax^2 + Bx + (C - y) = 0
A, B, C = constants from the ICAL quadratic regression
y = Response ratio = Area of analyte/Area of internal standard (IS)
x = Amount ratio = Concentration of analyte/concentration of IS

Step 2: Calculate y from Quantitation Report

y = Ax/Ais

Step 3: Solve for x using the quadratic formula
Ax^2 + Bx + C - y = 0

$$x = \frac{b \pm \sqrt{\left(b^2 - 4a(c - y)\right)}}{2a}$$ (Two possible solutions)

Step 4: Solve for analyte concentration Cx

Cx = (Cis)(Amount ratio)

Example Spreadsheet Calculation:

Value of A from plot:	**-0.00629**	
Value of B from plot:	**0.511**	
Value of C from plot:	**-0.0276**	
Area of unknown from quantitation report:	**293821**	
Area of IS from quantiation report:	**784848**	
Response ratio, y:	**0.374367**	
C - y:	**-0.40197**	
Root 1 - Computed amount ratio , X1:	**80.44567**	
Root 2 - Computed amount ratio , X2:	**0.794396**	use this solution
Concentration of IS, Cis:	**25.00**	
Concentration of analyte, Cx:	**19.86**	ug/L

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 051316
Analyst1: TMB Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 23
Method: 624 SOP: MSV10 Rev: 14
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID: 53752

Internal Standard: STD75899 Surrogate Standard: STD75899
CCV: STD76127 LCS: STD76109 MS/MSD: NA

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG568561

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
8M412209	WG568561-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 08:38
8M412210	WG568561-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 08:57
8M412211	WG568561-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 09:29
8M412212	WG568561-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 09:56
8M412213	WG568561-02 50ug/L CCV STD 8260	NA	1	1	STD76127	05/13/16 10:19
8M412214	WG568561-02 50ug/L CCV STD 8260	NA	1	1	STD76127	05/13/16 10:51
8M412215	WG568565-01 100ug/L A9 CCV STD 8260	NA	1	1	STD76072	05/13/16 11:36
8M412216	BFB CHECK	NA	1	1	STD76034	05/13/16 12:04
8M412217	BFB CHECK	NA	1	1	STD76034	05/13/16 12:18
8M412218	WG568561-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 12:36
8M412219	RINSE	NA	1	1		05/13/16 13:00
8M412220	WG568561-02 0.3ug/L STD 8260	NA	1	1	STD76127	05/13/16 13:29
8M412221	WG568561-03 0.4ug/L STD 8260	NA	1	1	STD76127	05/13/16 13:58
8M412222	WG568561-04 1ug/L STD 8260	NA	1	1	STD76127	05/13/16 14:27
8M412223	WG568561-05 2ug/L STD 8260	NA	1	1	STD76127	05/13/16 14:56
8M412224	WG568561-06 5ug/L STD 8260	NA	1	1	STD76127	05/13/16 15:25
8M412225	WG568561-07 20ug/L STD 8260	NA	1	1	STD76127	05/13/16 15:54
8M412226	WG568561-08 50ug/L STD 8260	NA	1	1	STD76127	05/13/16 16:23
8M412227	WG568561-09 100ug/L STD 8260	NA	1	1	STD76127	05/13/16 16:52
8M412228	WG568561-10 200ug/L STD 8260	NA	1	1	STD76127	05/13/16 17:21
8M412229	WG568561-11 300ug/L STD 8260	NA	1	1	STD76127	05/13/16 17:50
8M412230	RINSE	NA	1	1		05/13/16 18:19
8M412231	RINSE	NA	1	1		05/13/16 18:49
8M412232	WG568561-12 50ug/L ALT SRC STD 8260	NA	1	1	STD76109	05/13/16 19:17
8M412233	RINSE	NA	1	1		05/13/16 19:46

Comments

Seq.	Rerun	Dil.	Reason	Analytes
5	X			
File ID: 8M412213				
	Several compounds were low, DNR.			
6				
File ID: 8M412214				

Approved: May 20, 2016

Page: 1

Microbac Laboratories Inc.

Instrument Run Log

Instrument:	HPMS8	Dataset:	051316		
Analyst1:	TMB	Analyst2:	NA		
Method:	8260B	SOP:	MSV01	Rev:	23
Method:	624	SOP:	MSV10	Rev:	14
Method:	5030B/5030C/5035A	SOP:	PAT01	Rev:	18
Maintenance Log ID:	53752				

Internal Standard: STD75899 Surrogate Standard: STD75899

CCV: STD76127 LCS: STD76109 MS/MSD: NA

Column 1 ID: RTX502.2 Column 2 ID: NA

Workgroups: WG568561

Comments:

Comments

Seq.	Rerun	Dil.	Reason	Analytes
			Two C compounds failed. DNR. Run a curve.	
7 File ID: 8M412215				
			DNR.	

Approved: May 20, 2016

Page: 2

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 062116
Analyst1: FJB Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 23
Method: OVAP SOP: MSV01/PAT01 Rev: 0
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID: 53790

Internal Standard: STD75370 Surrogate Standard: STD76370
CCV: STD76578 LCS: STD76597 MS/MSD: STD76597

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573645

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
8M413124	WG573644-01 BFB 50ng 8260	NA	1	1	STD76496	06/21/16 17:23
8M413125	WG573644-02 50ug/L CCV 8260	NA	1	1	STD76715	06/21/16 17:51
8M413126	WG573644-02 50ug/L CCV 8260	NA	1	1	STD76578	06/21/16 18:23
8M413127	WG573787-01 100ug/L CCV A9	NA	1	1	STDXXXXX	06/21/16 18:52
8M413128	WG573645-01 BLANK 8260	NA	1	1		06/21/16 19:21
8M413129	WG573645-02 20ug/L LCS 8260	NA	1	1	STD76744	06/21/16 19:50
8M413130	WG573645-03 20ug/L LCS2 8260	NA	1	1	STD76744	06/21/16 20:19
8M413131	L16060994-11 5X B 826-SPE D1	<2	1	5		06/21/16 20:48
8M413132	L16060724-01 5X B 8260 D1	<2	1	5		06/21/16 21:17
8M413133	L16061128-04 A 826-SPE1	5	1	1		06/21/16 21:46
8M413134	L16061128-05 A 826-SPE1	5	1	1		06/21/16 22:15
8M413135	L16060927-05 A 826-LOW	<2	1	1		06/21/16 22:44
8M413136	L16061016-01 A 826-SPE	<2	1	1		06/21/16 23:14
8M413137	L16061016-20 A 826-SPE	<2	1	1		06/21/16 23:43
8M413138	L16061016-21 A 826-SPE	<2	1	1		06/22/16 00:12
8M413139	L16060994-09 B 826-SPE	<2	1	1		06/22/16 00:41
8M413140	L16060998-01 5X B 826-SPE 00	<2	1	5		06/22/16 01:10
8M413141	L16061128-01 A 826-SPE1	7	1	1		06/22/16 01:38
8M413142	L16061128-02 A 826-SPE1	7	1	1		06/22/16 02:07
8M413143	L16061128-03 A 826-SPE1	7	1	1		06/22/16 02:35
8M413144	L16061001-01 A 826-SPE	<2	1	1		06/22/16 03:05
8M413145	L16060927-04 A 826-LOW	<2	1	1		06/22/16 03:34
8M413146	L16060927-01 A 826-LOW	<2	1	1		06/22/16 04:03
8M413147	L16060927-03 A 826-LOW	<2	1	1		06/22/16 04:32
8M413148	L16060927-02 20X A 826-LOW	<2	1	20		06/22/16 05:00
8M413149	WG573648-01 20ug/L LCS 624	NA	2	1	STD76744	06/22/16 05:30
8M413150	WG573648-02 20ug/L LCS2 624	NA	2	1	STD76744	06/22/16 05:59
8M413151	RINSE	NA	2	1		06/22/16 06:28
8M413152	WG573648-03 BLANK 624	NA	2	1		06/22/16 06:57
8M413153	L16061067-01 A 624-SPE	6	2	1		06/22/16 07:26
8M413154	L16061079-01 A 624-SPE5	6	2	1		06/22/16 07:55
8M413155	L16061134-02 A 624-SPE	8	2	1		06/22/16 08:23
8M413156	L16061135-02 A 624-SPE	7	2	1		06/22/16 08:52
8M413157	L16061170-06 A 624-SPE1	<2	2	1		06/22/16 09:22

Approved: June 30, 2016

Page: 1

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 062116
Analyst1: FJB Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 23
Method: OVAP SOP: MSV01/PAT01 Rev: 0
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID: 53790

Internal Standard: STD75370 Surrogate Standard: STD76370
CCV: STD76578 LCS: STD76597 MS/MSD: STD76597

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573645

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
8M413158	L16061132-01 50X A 624-SPE1	7	2	50		06/22/16 09:51
8M413159	L16061123-01 10X A 624-SPE	7	2	10		06/22/16 10:20
8M413160	L16061123-02 100X A 624-SPE	7	2	100		06/22/16 10:49

Comments

Seq.	Rerun	Dil.	Reason	Analytes
23	X	200	Over Calibration Range	tce
File ID: 8M413146				
L16060927-01 sample labelled incorrectly, was actually 02.				
24	X	1	Carry-over contamination	
File ID: 8M413147				
L16060927-03				
25	X	1	Analyzed too dilute	
File ID: 8M413148				
L16060927-02 vial labelled incorrectly, was actually sample 01				
30				
File ID: 8M413153				
L16061067-01 CT1 Cooler temp out				
35				
File ID: 8M413158				
L16061132-01 Reporting limits elevated due to foamy sample matrix. Attempt at 10X dilution with antifoam was aborted by instrument foam sensor.				

Approved: June 30, 2016

Page: 2

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 062216
Analyst1: FJB Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 23
Method: OVAP SOP: MSV01/PAT01 Rev: 0
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID:

Internal Standard: STD75370 Surrogate Standard: STD76370
CCV: STD76578 LCS: STD76744 MS/MSD: STD76744

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573771

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
8M413161	WG573770-01 BFB 50ng 8260	NA	1	1	STD76496	06/22/16 14:38
8M413162	WG573770-02 50ug/L CCV 8260	NA	1	1	STD76578	06/22/16 15:06
8M413163	WG573770-02 50ug/L CCV 8260	NA	1	1	STD76578	06/22/16 15:41
8M413164	WGXXXXXX-01 100ug/L A9CCV 8260	NA	1	1	STDXXXXX	06/22/16 16:10
8M413165	WG573771-01 BLANK 8260	NA	1	1		06/22/16 16:39
8M413166	WG573771-02 20ug/L LCS 8260	NA	1	1	STD76744	06/22/16 17:13
8M413167	L16060829-02 10X 826-TC-SPE	NA	17	10		06/22/16 17:42
8M413168	L16060927-01 200X B 826-LOW	<2	1	200		06/22/16 18:11
8M413169	L16060927-02 B 826-LOW	<2	1	1		06/22/16 18:40
8M413170	L16060927-03 B 826-LOW	<2	1	1		06/22/16 19:09
8M413171	L16060896-06 TB A 826-SPE	<2	1	1		06/22/16 19:38
8M413172	L16060896-07 RB A 826-SPE	<2	1	1		06/22/16 20:07
8M413173	L16060930-04 TB A 826-SPE	<2	1	1		06/22/16 20:37
8M413174	L16060986-21 REF A 826-SPE	<2	1	1		06/22/16 21:06
8M413175	L16060986-31 MS A 826-SPE	<2	1	1	STD76744	06/22/16 21:35
8M413176	L16060986-33 MSD A 826-SPE	<2	1	1	STD76744	06/22/16 22:03
8M413177	L16060986-01 A 826-SPE	<2	1	1		06/22/16 22:32
8M413178	L16060986-03 A 826-SPE	<2	1	1		06/22/16 23:01
8M413179	L16060986-05 A 826-SPE	<2	1	1		06/22/16 23:30
8M413180	L16060896-01 A 826-SPE	<2	1	1		06/22/16 23:59
8M413181	L16060896-02 A 826-SPE	<2	1	1		06/23/16 00:27
8M413182	L16060896-03 A 826-SPE	<2	1	1		06/23/16 00:56
8M413183	L16060896-04 A 826-SPE	<2	1	1		06/23/16 01:25
8M413184	L16060896-05 A 826-SPE	<2	1	1		06/23/16 01:54
8M413185	L16060930-03 10X A 826-SPE	7	1	10		06/23/16 02:23
8M413186	CCV	NA	1	1		06/23/16 02:51
8M413187	RINSE	NA	1	1		06/23/16 03:21
8M413188	WG573023-01 10X FBLK	NA	17	10		06/23/16 03:50
8M413189	10X GROUNDWATER	NA	1	10		06/23/16 04:19

Comments

Seq.	Rerun	Dil.	Reason	Analytes
8	X	1	Analyzed too dilute	

Approved: June 27, 2016

Sarah Vandenberg

Page: 1

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 062216
Analyst1: FJB Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 23
Method: OVAP SOP: MSV01/PAT01 Rev: 0
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID:

Internal Standard: STD75370 Surrogate Standard: STD76370
CCV: STD76578 LCS: STD76744 MS/MSD: STD76744

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573771

Comments:

Comments

Seq.	Rerun	Dil.	Reason	Analytes
File ID: 8M413168				
	L16060927-01		Samples 01 and 02 A and C vials labels had been switched. This caused B vials to be analyzed at incorrect dilutions.	
9	X	200	Over Calibration Range	tce
File ID: 8M413169				
	L16060927-02		Samples 01 and 02 A and C vials labels had been switched. This caused B vials to be analyzed at incorrect dilutions.	
10	X	1	Carry-over contamination	
File ID: 8M413170				
	L16060927-03			
11	X	1	Carry-over contamination	
File ID: 8M413171				
	L16060896-06			
12	X	1	Carry-over contamination	
File ID: 8M413172				
	L16060896-07			
25	X	1	Analyzed too dilute	
File ID: 8M413185				
	L16060930-03			

Approved: June 27, 2016

Sarah Vandenberg

Page: 2

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 062316
Analyst1: FJB Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 23
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18

Maintenance Log ID:

Internal Standard: STD76370 Surrogate Standard: STD76370
CCV: STD76715 LCS: STD76744 MS/MSD: STD76744

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573991

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
8M413190	WG573990-01 BFB 50ng 8260	NA	1	1	STD76496	06/23/16 19:56
8M413191	WG573990-02 50ug/L CCV 8260	NA	1	1	STD76715	06/23/16 20:24
8M413192	WG573990-02 50ug/L CCV 8260	NA	1	1	STD76715	06/23/16 21:00
8M413193	WG573XXX-01 100ug/L CCV A9	NA	1	1	STDXXXXX	06/23/16 21:30
8M413194	WG573991-01 BLANK 8260	NA	1	1		06/23/16 21:59
8M413195	WG573991-02 20ug/L LCS 8260	NA	1	1	STD76744	06/23/16 22:28
8M413196	L16060927-02 200X C 826-LOW D1	<2	1	200		06/23/16 22:56
8M413197	L16060978-05 A 826-SPE	<2	1	1		06/23/16 23:26
8M413198	L16060992-01 A 826-SPE	<2	1	1		06/23/16 23:55
8M413199	L16060927-01 C 826-LOW	<2	1	1		06/24/16 00:24
8M413200	L16060927-03 C 826-LOW	<2	1	1		06/24/16 00:53
8M413201	L16060992-04 A 826-SPE	<2	1	1		06/24/16 01:22
8M413202	L16061095-12 A 826-SPE	<2	1	1		06/24/16 01:51
8M413203	L16061095-13 A 826-SPE	<2	1	1		06/24/16 02:20
8M413204	L16060896-06 B 826-SPE	<2	1	1		06/24/16 02:49
8M413205	L16060896-07 B 826-SPE	<2	1	1		06/24/16 03:18
8M413206	L16060930-03 B+C 826-SPE	7	1	1	SEDIMENT	06/24/16 03:47
8M413207	L16060730-03 B 826-LOW A1	<2	1	1		06/24/16 04:16
8M413208	L16060823-04 B 826-LOW A1	<2	1	1		06/24/16 04:46
8M413209	L16061068-01 REF 10X A 826-LOW	NA	1	10		06/24/16 05:15
8M413210	L16061068-02 MS 10X A 826-LOW	NA	1	10	STD76744	06/24/16 05:44
8M413211	L16061068-03 MSD 10X A 826-LOW	NA	1	10	STD76744	06/24/16 06:13
8M413212	L16060978-03 10X A 826-SPE	11	1	10		06/24/16 06:42
8M413213	L16060978-04 10X A 826-SPE	11	1	10		06/24/16 07:11
8M413214	L16061068-04 A 826-LOW	<2	1	1		06/24/16 07:40
8M413215	RINSE	NA	1	1		06/24/16 08:09
8M413216	RINSE	NA	1	1		06/24/16 08:38
8M413217	GROUNDWATER	NA	1	1		06/24/16 09:07

Comments

Seq.	Rerun	Dil.	Reason	Analytes
2	X			
File ID: 8M413191				

Approved: June 29, 2016

Sarah Vandenberg

Page: 1

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS8 Dataset: 062316
Analyst1: FJB Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 23
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18

Maintenance Log ID:

Internal Standard: STD76370 Surrogate Standard: STD76370
CCV: STD76715 LCS: STD76744 MS/MSD: STD76744

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573991

Comments:

Comments

Seq.	Rerun	Dil.	Reason	Analytes
DNR.				
4				
File ID: 8M413193				
	Not needed, DNR.			
23	X	100	Over Calibration Range	TOLUENE
File ID: 8M413212				
25			Carry-over contamination	
File ID: 8M413214				

Approved: June 29, 2016

Sarah Vandenberg

Page: 2

Microbac Laboratories Inc.

Data Checklist

Date: 13-MAY-2016
Analyst: TMB
Analyst: NA
Method: 8260B/624
Instrument: HPMS8
Curve Workgroup: NA
Runlog ID: 75126
Analytical Workgroups: WG568561

System Performance Check	NA
BFB	X
Initial Calibration	X
Average RF	X
Linear Reg or Higher Order Curve	X
Second Source standard % Difference	X
Continuing Calibration /Check Standards	X
Project/Client Specific Requirements	X
Special Standards	X
Blanks	X
TCL's	X
Surrogates	X
LCS (Laboratory Control Sample)	X
Recoveries	X
Surrogates	X
MS/MSD/Duplicates	NA
Samples	X
TCL Hits	X
Spectra of TCL Hits	X
Surrogates	TMB
Internal Standards Criteria	X
Library Searches	NA
Calculations & Correct Factors	X
Dilutions Run	NA
Reruns	NA
Manual Integrations	NA
Case Narrative	X
Results Reporting/Data Qualifiers	X
KOBRA Workgroup Data	X
Check for Completeness	X
Primary Reviewer	TMB
Secondary Reviewer	FJB
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Check the reasonableness of the results	X

Primary Reviewer:
17-MAY-2016
Tiffany Bailey

Secondary Reviewer:
20-MAY-2016

CHECKLIST1 - Modified 03/05/2008
Generated: MAY-20-2016 11:18:36

Microbac Laboratories Inc.

Data Checklist

Date: 21-JUN-2016
Analyst: FJB
Analyst: NA
Method: 8260B/624
Instrument: HPMS8
Curve Workgroup: NA
Runlog ID: 75920
Analytical Workgroups: WG573645, WG573648

System Performance Check	X
BFB	X
Initial Calibration	X
Average RF	X
Linear Reg or Higher Order Curve	X
Second Source standard % Difference	X
Continuing Calibration /Check Standards	X
Project/Client Specific Requirements	X
Special Standards	NA
Blanks	X
TCL's	X
Surrogates	X
LCS (Laboratory Control Sample)	X
Recoveries	X
Surrogates	X
MS/MSD/Duplicates	NA
Samples	X
TCL Hits	X
Spectra of TCL Hits	TMB
Surrogates	X
Internal Standards Criteria	X
Library Searches	NA
Calculations & Correct Factors	X
Dilutions Run	X
Reruns	X
Manual Integrations	NA
Case Narrative	X
Results Reporting/Data Qualifiers	X
KOBRA Workgroup Data	X
Check for Completeness	X
Primary Reviewer	FJB
Secondary Reviewer	WTD
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Check the reasonableness of the results	X

Primary Reviewer:
30-JUN-2016

Secondary Reviewer:
30-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-30-2016 14:47:57

Microbac Laboratories Inc.

Data Checklist

Date: 22-JUN-2016
Analyst: FJB
Analyst: NA
Method: 8260B
Instrument: HPMS8
Curve Workgroup: NA
Runlog ID: 75952
Analytical Workgroups: WG573771

System Performance Check	X
BFB	X
Initial Calibration	X
Average RF	X
Linear Reg or Higher Order Curve	X
Second Source standard % Difference	X
Continuing Calibration /Check Standards	X
Project/Client Specific Requirements	X
Special Standards	NA
Blanks	X
TCL's	X
Surrogates	X
LCS (Laboratory Control Sample)	X
Recoveries	X
Surrogates	X
MS/MSD/Duplicates	X
Samples	X
TCL Hits	X
Spectra of TCL Hits	WTD
Surrogates	X
Internal Standards Criteria	X
Library Searches	NA
Calculations & Correct Factors	X
Dilutions Run	X
Reruns	X
Manual Integrations	NA
Case Narrative	X
Results Reporting/Data Qualifiers	X
KOBRA Workgroup Data	X
Check for Completeness	X
Primary Reviewer	FJB
Secondary Reviewer	SAV
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Check the reasonableness of the results	X

Primary Reviewer:
27-JUN-2016

Secondary Reviewer:
27-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-27-2016 16:24:53

Microbac Laboratories Inc.

Data Checklist

Date: 23-JUN-2016
Analyst: FJB
Analyst: NA
Method: 8260B
Instrument: HPMS8
Curve Workgroup: NA
Runlog ID: 76008
Analytical Workgroups: WG573991

System Performance Check	NA
BFB	X
Initial Calibration	X
Average RF	X
Linear Reg or Higher Order Curve	X
Second Source standard % Difference	X
Continuing Calibration /Check Standards	X
Project/Client Specific Requirements	X
Special Standards	NA
Blanks	X
TCL's	X
Surrogates	X
LCS (Laboratory Control Sample)	X
Recoveries	X
Surrogates	X
MS/MSD/Duplicates	X
Samples	X
TCL Hits	X
Spectra of TCL Hits	WTD
Surrogates	X
Internal Standards Criteria	X
Library Searches	NA
Calculations & Correct Factors	X
Dilutions Run	X
Reruns	X
Manual Integrations	NA
Case Narrative	X
Results Reporting/Data Qualifiers	X
KOBRA Workgroup Data	X
Check for Completeness	X
Primary Reviewer	TMB
Secondary Reviewer	SAV
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Check the reasonableness of the results	X

Primary Reviewer:
29-JUN-2016
Tiffany Bailey

Secondary Reviewer:
29-JUN-2016
Sarah Vandenberg

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-29-2016 14:37:11

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 8260B

AAB#: WG573645

Login Number: L16060927

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18CPTMW19-061516	04	06/15/16					06/22/2016	6.7	14		06/22/16	6.7	14	
TRIP BLANK	05	06/15/16					06/21/2016	6.9	14		06/21/16	6.9	14	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4831243
Report generated 06/30/2016 15:02

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 8260B

AAB#: WG573771

Login Number: L16060927

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18CPTMW23-061516	02	06/15/16					06/22/2016	7.4	14		06/22/16	7.4	14	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4831243
Report generated 06/30/2016 15:02

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 8260B

AAB#: WG573991

Login Number: L16060927

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18CPTMW23SW-061516	01	06/15/16					06/24/2016	8.7	14		06/24/16	8.7	14	
18CPTMW23-061516	02	06/15/16					06/23/2016	8.6	14		06/23/16	8.6	14	
18CPTMW19SW-061516	03	06/15/16					06/24/2016	8.6	14		06/24/16	8.6	14	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4831243
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.
SURROGATE STANDARDS

Login Number: L16060927
Instrument Id: HPMS8
Workgroup (AAB#): WG573645

Method: 8260
CAL ID: HPMS8-13-MAY-16
Matrix: Water

Sample Number	Dilution	Tag	1	2	3	4
L16060927-04	1.00	01	114	112	99.9	104
L16060927-05	1.00	01	112	111	100	103
WG573645-01	1.00	01	110	111	99.8	105
WG573645-02	1.00	01	109	107	98.0	103
WG573645-03	1.00	01	106	108	99.0	102

Surrogates	Surrogate Limits		
1 - 1,2-Dichloroethane-d4	70	-	120
2 - Dibromofluoromethane	85	-	115
3 - 4-Bromofluorobenzene	75	-	120
4 - Toluene-d8	85	-	120

Underline = Result out of surrogate limits

DL = surrogate diluted out
ND = surrogate not detected

SURROGATES - Modified 03/06/2008
PDF File ID: 4831249
Report generated: 06/30/2016 15:03

Microbac Laboratories Inc.
SURROGATE STANDARDS

Login Number: L16060927
Instrument Id: HPMS8
Workgroup (AAB#): WG573991

Method: 8260
CAL ID: HPMS8-13-MAY-16
Matrix: Water

Sample Number	Dilution	Tag	1	2	3	4
L16060927-01	1.00	01	116	111	101	104
L16060927-02	200	DL01	116	112	101	101
L16060927-03	1.00	01	119	113	103	104
WG573991-01	1.00	01	115	113	102	102
WG573991-02	1.00	01	113	110	98.9	103

Surrogates	Surrogate Limits		
1 - 1,2-Dichloroethane-d4	70	-	120
2 - Dibromofluoromethane	85	-	115
3 - 4-Bromofluorobenzene	75	-	120
4 - Toluene-d8	85	-	120

Underline = Result out of surrogate limits

DL = surrogate diluted out
ND = surrogate not detected

SURROGATES - Modified 03/06/2008
PDF File ID: 4831249
Report generated: 06/30/2016 15:03

Microbac Laboratories Inc.
SURROGATE STANDARDS

Login Number: L16060927
Instrument Id: HPMS8
Workgroup (AAB#): WG573771

Method: 8260
CAL ID: HPMS8-13-MAY-16
Matrix: Water

Sample Number	Dilution	Tag	1	2	3	4
L16060927-02	1.00	01	115	112	99.4	105
WG573771-01	1.00	01	111	107	98.4	104
WG573771-02	1.00	01	113	110	103	103

Surrogates	Surrogate Limits		
1 - 1,2-Dichloroethane-d4	70	-	120
2 - Dibromofluoromethane	85	-	115
3 - 4-Bromofluorobenzene	75	-	120
4 - Toluene-d8	85	-	120

Underline = Result out of surrogate limits

DL = surrogate diluted out
ND = surrogate not detected

SURROGATES - Modified 03/06/2008
PDF File ID: 4831249
Report generated: 06/30/2016 15:03

METHOD BLANK SUMMARY

Login Number: L16060927
Blank File ID: 8M413128
Prep Date: 06/21/16 19:21
Analyzed Date: 06/21/16 19:21
Analyst: FJB

Work Group: WG573645
Blank Sample ID: WG573645-01
Instrument ID: HPMS8
Method: 8260B

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573645-02	8M413129	06/21/16 19:50	01
LCS2	WG573645-03	8M413130	06/21/16 20:19	01
TRIP BLANK	L16060927-05	8M413135	06/21/16 22:44	01
18CPTMW19-061516	L16060927-04	8M413145	06/22/16 03:34	01

Report Name: BLANK_SUMMARY
PDF File ID: 4833866
Report generated 06/30/2016 15:02

METHOD BLANK SUMMARY

Login Number: L16060927
Blank File ID: 8M413194
Prep Date: 06/23/16 21:59
Analyzed Date: 06/23/16 21:59
Analyst: FJB

Work Group: WG573991
Blank Sample ID: WG573991-01
Instrument ID: HPMS8
Method: 8260B

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573991-02	8M413195	06/23/16 22:28	01
18CPTMW23-061516	L16060927-02	8M413196	06/23/16 22:56	DL01
18CPTMW23SW-061516	L16060927-01	8M413199	06/24/16 00:24	01
18CPTMW19SW-061516	L16060927-03	8M413200	06/24/16 00:53	01

Report Name: BLANK_SUMMARY
PDF File ID: 4833866
Report generated 06/30/2016 15:02

METHOD BLANK SUMMARY

Login Number: L16060927
Blank File ID: 8M413165
Prep Date: 06/22/16 16:39
Analyzed Date: 06/22/16 16:39
Analyst: FJB

Work Group: WG573771
Blank Sample ID: WG573771-01
Instrument ID: HPMS8
Method: 8260B

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573771-02	8M413166	06/22/16 17:13	01
18CPTMW23-061516	L16060927-02	8M413169	06/22/16 18:40	01

Report Name: BLANK_SUMMARY
PDF File ID: 4833866
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060927 Prep Date: 06/21/16 19:21 Sample ID: WG573645-01
Instrument ID: HPMS8 Run Date: 06/21/16 19:21 Prep Method: 5030B/5030C/503
File ID: 8M413128 Analyst: FJB Method: 8260B
Workgroup (AAB#): WG573645 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS8-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Acetone	2.50	10.0	2.50	1	U
Benzene	0.125	1.00	0.125	1	U
Bromobenzene	0.125	1.00	0.125	1	U
Bromochloromethane	0.200	1.00	0.200	1	U
Bromodichloromethane	0.250	1.00	0.250	1	U
Bromoform	0.500	2.00	0.500	1	U
Bromomethane	0.500	2.00	0.500	1	U
2-Butanone	2.50	10.0	2.50	1	U
n-Butylbenzene	0.250	1.00	0.250	1	U
sec-Butylbenzene	0.250	1.00	0.250	1	U
tert-Butylbenzene	0.250	1.00	0.250	1	U
Carbon disulfide	0.500	2.00	0.500	1	U
Carbon tetrachloride	0.250	1.00	0.250	1	U
Chlorobenzene	0.125	1.00	0.125	1	U
Chlorodibromomethane	0.250	1.00	0.250	1	U
Chloroethane	0.500	2.00	0.500	1	U
Chloroform	0.125	1.00	0.125	1	U
Chloromethane	0.500	2.00	0.500	1	U
2-Chlorotoluene	0.125	1.00	0.125	1	U
4-Chlorotoluene	0.250	1.00	0.250	1	U
1,2-Dibromo-3-chloropropane	1.00	5.00	1.00	1	U
1,2-Dibromoethane	0.250	1.00	0.250	1	U
Dibromomethane	0.250	1.00	0.250	1	U
1,2-Dichlorobenzene	0.125	1.00	0.125	1	U
1,3-Dichlorobenzene	0.250	1.00	0.250	1	U
1,4-Dichlorobenzene	0.125	1.00	0.125	1	U
Dichlorodifluoromethane	0.250	1.00	0.250	1	U
1,1-Dichloroethane	0.125	1.00	0.125	1	U
1,2-Dichloroethane	0.250	1.00	0.250	1	U
1,1-Dichloroethene	0.500	2.00	0.500	1	U
cis-1,2-Dichloroethene	0.250	1.00	0.250	1	U
trans-1,2-Dichloroethene	0.250	1.00	0.250	1	U
1,2-Dichloropropane	0.200	1.00	0.200	1	U
1,3-Dichloropropane	0.200	1.00	0.200	1	U
2,2-Dichloropropane	0.250	1.00	0.250	1	U
cis-1,3-Dichloropropene	0.250	1.00	0.250	1	U
trans-1,3-Dichloropropene	0.500	2.00	0.500	1	U
1,1-Dichloropropene	0.250	1.00	0.250	1	U
Ethylbenzene	0.250	1.00	0.250	1	U
2-Hexanone	2.50	10.0	2.50	1	U
Hexachlorobutadiene	0.250	1.00	0.250	1	U
Isopropylbenzene	0.250	1.00	0.250	1	U

Report Name:BLANK
PDF ID: 4831244
30-JUN-2016 15:02

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060927 Prep Date: 06/21/16 19:21 Sample ID: WG573645-01
Instrument ID: HPMS8 Run Date: 06/21/16 19:21 Prep Method: 5030B/5030C/503
File ID: 8M413128 Analyst: FJB Method: 8260B
Workgroup (AAB#): WG573645 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS8-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
p-Isopropyltoluene	0.250	1.00	0.250	1	U
4-Methyl-2-pentanone	2.50	10.0	2.50	1	U
Methylene chloride	0.250	1.00	0.250	1	U
Naphthalene	0.200	1.00	0.200	1	U
n-Propylbenzene	0.125	1.00	0.125	1	U
Styrene	0.125	1.00	0.125	1	U
1,1,1,2-Tetrachloroethane	0.250	1.00	0.250	1	U
1,1,2,2-Tetrachloroethane	0.200	1.00	0.200	1	U
Tetrachloroethene	0.250	1.00	0.250	1	U
Toluene	0.250	1.00	0.250	1	U
1,2,3-Trichlorobenzene	0.150	1.00	0.150	1	U
1,2,4-Trichlorobenzene	0.200	1.00	0.200	1	U
1,1,1-Trichloroethane	0.250	1.00	0.250	1	U
1,1,2-Trichloroethane	0.250	1.00	0.250	1	U
Trichloroethene	0.250	1.00	0.250	1	U
Trichlorofluoromethane	0.250	1.00	0.250	1	U
1,2,3-Trichloropropane	0.500	2.00	0.500	1	U
1,2,4-Trimethylbenzene	0.250	1.00	0.250	1	U
1,3,5-Trimethylbenzene	0.250	1.00	0.250	1	U
Vinyl chloride	0.250	1.00	0.250	1	U
o-Xylene	0.250	1.00	0.250	1	U
m-,p-Xylene	0.500	2.00	0.500	1	U

Surrogates	% Recovery	Surrogate Limits	Qualifier
Dibromofluoromethane	111	85 - 115	PASS
1,2-Dichloroethane-d4	110	70 - 120	PASS
Toluene-d8	105	85 - 120	PASS
4-Bromofluorobenzene	99.8	75 - 120	PASS

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4831244
30-JUN-2016 15:02

Microbac Laboratories Inc.

METHOD BLANK REPORT

Login Number: L16060927 Prep Date: 06/22/16 16:39 Sample ID: WG573771-01
Instrument ID: HPMS8 Run Date: 06/22/16 16:39 Prep Method: 5030B/5030C/503
File ID: 8M413165 Analyst: FJB Method: 8260B
Workgroup (AAB#): WG573771 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS8-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Acetone	2.50	10.0	2.50	1	U
Benzene	0.125	1.00	0.125	1	U
Bromobenzene	0.125	1.00	0.125	1	U
Bromochloromethane	0.200	1.00	0.200	1	U
Bromodichloromethane	0.250	1.00	0.250	1	U
Bromoform	0.500	2.00	0.500	1	U
Bromomethane	0.500	2.00	0.500	1	U
2-Butanone	2.50	10.0	2.50	1	U
n-Butylbenzene	0.250	1.00	0.250	1	U
sec-Butylbenzene	0.250	1.00	0.250	1	U
tert-Butylbenzene	0.250	1.00	0.250	1	U
Carbon disulfide	0.500	2.00	0.500	1	U
Carbon tetrachloride	0.250	1.00	0.250	1	U
Chlorobenzene	0.125	1.00	0.125	1	U
Chlorodibromomethane	0.250	1.00	0.250	1	U
Chloroethane	0.500	2.00	0.500	1	U
Chloroform	0.125	1.00	0.125	1	U
Chloromethane	0.500	2.00	0.500	1	U
2-Chlorotoluene	0.125	1.00	0.125	1	U
4-Chlorotoluene	0.250	1.00	0.250	1	U
1,2-Dibromo-3-chloropropane	1.00	5.00	1.00	1	U
1,2-Dibromoethane	0.250	1.00	0.250	1	U
Dibromomethane	0.250	1.00	0.250	1	U
1,2-Dichlorobenzene	0.125	1.00	0.125	1	U
1,3-Dichlorobenzene	0.250	1.00	0.250	1	U
1,4-Dichlorobenzene	0.125	1.00	0.125	1	U
Dichlorodifluoromethane	0.250	1.00	0.250	1	U
1,1-Dichloroethane	0.125	1.00	0.125	1	U
1,2-Dichloroethane	0.250	1.00	0.250	1	U
1,1-Dichloroethene	0.500	2.00	0.500	1	U
cis-1,2-Dichloroethene	0.250	1.00	0.250	1	U
trans-1,2-Dichloroethene	0.250	1.00	0.250	1	U
1,2-Dichloropropane	0.200	1.00	0.200	1	U
1,3-Dichloropropane	0.200	1.00	0.200	1	U
2,2-Dichloropropane	0.250	1.00	0.250	1	U
cis-1,3-Dichloropropene	0.250	1.00	0.250	1	U
trans-1,3-Dichloropropene	0.500	2.00	0.500	1	U
1,1-Dichloropropene	0.250	1.00	0.250	1	U
Ethylbenzene	0.250	1.00	0.250	1	U
2-Hexanone	2.50	10.0	2.50	1	U
Hexachlorobutadiene	0.250	1.00	0.250	1	U
Isopropylbenzene	0.250	1.00	0.250	1	U

Report Name: BLANK
PDF ID: 4831244
30-JUN-2016 15:02

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060927 Prep Date: 06/22/16 16:39 Sample ID: WG573771-01
Instrument ID: HPMS8 Run Date: 06/22/16 16:39 Prep Method: 5030B/5030C/503
File ID: 8M413165 Analyst: FJB Method: 8260B
Workgroup (AAB#): WG573771 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS8-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
p-Isopropyltoluene	0.250	1.00	0.250	1	U
4-Methyl-2-pentanone	2.50	10.0	2.50	1	U
Methylene chloride	0.250	1.00	0.250	1	U
Naphthalene	0.200	1.00	0.200	1	U
n-Propylbenzene	0.125	1.00	0.125	1	U
Styrene	0.125	1.00	0.125	1	U
1,1,1,2-Tetrachloroethane	0.250	1.00	0.250	1	U
1,1,2,2-Tetrachloroethane	0.200	1.00	0.200	1	U
Tetrachloroethene	0.250	1.00	0.250	1	U
Toluene	0.250	1.00	0.250	1	U
1,2,3-Trichlorobenzene	0.150	1.00	0.150	1	U
1,2,4-Trichlorobenzene	0.200	1.00	0.200	1	U
1,1,1-Trichloroethane	0.250	1.00	0.250	1	U
1,1,2-Trichloroethane	0.250	1.00	0.250	1	U
Trichlorofluoromethane	0.250	1.00	0.250	1	U
1,2,3-Trichloropropane	0.500	2.00	0.500	1	U
1,2,4-Trimethylbenzene	0.250	1.00	0.250	1	U
1,3,5-Trimethylbenzene	0.250	1.00	0.250	1	U
Vinyl chloride	0.250	1.00	0.250	1	U
o-Xylene	0.250	1.00	0.250	1	U
m-,p-Xylene	0.500	2.00	0.500	1	U

Surrogates	% Recovery	Surrogate Limits	Qualifier
Dibromofluoromethane	107	85 - 115	PASS
1,2-Dichloroethane-d4	111	70 - 120	PASS
Toluene-d8	104	85 - 120	PASS
4-Bromofluorobenzene	98.4	75 - 120	PASS

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4831244
30-JUN-2016 15:02

Microbac Laboratories Inc.

METHOD BLANK REPORT

Login Number: L16060927 Prep Date: 06/23/16 21:59 Sample ID: WG573991-01
Instrument ID: HPMS8 Run Date: 06/23/16 21:59 Prep Method: 5030B/5030C/503
File ID: 8M413194 Analyst: FJB Method: 8260B
Workgroup (AAB#): WG573991 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS8-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Acetone	2.50	10.0	2.50	1	U
Benzene	0.125	1.00	0.125	1	U
Bromobenzene	0.125	1.00	0.125	1	U
Bromochloromethane	0.200	1.00	0.200	1	U
Bromodichloromethane	0.250	1.00	0.250	1	U
Bromoform	0.500	2.00	0.500	1	U
Bromomethane	0.500	2.00	0.500	1	U
2-Butanone	2.50	10.0	2.50	1	U
n-Butylbenzene	0.250	1.00	0.250	1	U
sec-Butylbenzene	0.250	1.00	0.250	1	U
tert-Butylbenzene	0.250	1.00	0.250	1	U
Carbon disulfide	0.500	2.00	0.500	1	U
Carbon tetrachloride	0.250	1.00	0.250	1	U
Chlorobenzene	0.125	1.00	0.125	1	U
Chlorodibromomethane	0.250	1.00	0.250	1	U
Chloroethane	0.500	2.00	0.500	1	U
Chloroform	0.125	1.00	0.125	1	U
Chloromethane	0.500	2.00	0.500	1	U
2-Chlorotoluene	0.125	1.00	0.125	1	U
4-Chlorotoluene	0.250	1.00	0.250	1	U
1,2-Dibromo-3-chloropropane	1.00	5.00	1.00	1	U
1,2-Dibromoethane	0.250	1.00	0.250	1	U
Dibromomethane	0.250	1.00	0.250	1	U
1,2-Dichlorobenzene	0.125	1.00	0.125	1	U
1,3-Dichlorobenzene	0.250	1.00	0.250	1	U
1,4-Dichlorobenzene	0.125	1.00	0.125	1	U
Dichlorodifluoromethane	0.250	1.00	0.250	1	U
1,1-Dichloroethane	0.125	1.00	0.125	1	U
1,2-Dichloroethane	0.250	1.00	0.250	1	U
1,1-Dichloroethene	0.500	2.00	0.500	1	U
cis-1,2-Dichloroethene	0.250	1.00	0.250	1	U
trans-1,2-Dichloroethene	0.250	1.00	0.250	1	U
1,2-Dichloropropane	0.200	1.00	0.200	1	U
1,3-Dichloropropane	0.200	1.00	0.200	1	U
2,2-Dichloropropane	0.250	1.00	0.250	1	U
cis-1,3-Dichloropropene	0.250	1.00	0.250	1	U
trans-1,3-Dichloropropene	0.500	2.00	0.500	1	U
1,1-Dichloropropene	0.250	1.00	0.250	1	U
Ethylbenzene	0.250	1.00	0.250	1	U
2-Hexanone	2.50	10.0	2.50	1	U
Hexachlorobutadiene	0.250	1.00	0.250	1	U
Isopropylbenzene	0.250	1.00	0.250	1	U

Report Name: BLANK
PDF ID: 4831244
30-JUN-2016 15:02

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060927 Prep Date: 06/23/16 21:59 Sample ID: WG573991-01
Instrument ID: HPMS8 Run Date: 06/23/16 21:59 Prep Method: 5030B/5030C/503
File ID: 8M413194 Analyst: FJB Method: 8260B
Workgroup (AAB#): WG573991 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS8-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
p-Isopropyltoluene	0.250	1.00	0.250	1	U
4-Methyl-2-pentanone	2.50	10.0	2.50	1	U
Methylene chloride	0.250	1.00	0.250	1	U
Naphthalene	0.200	1.00	0.200	1	U
n-Propylbenzene	0.125	1.00	0.125	1	U
Styrene	0.125	1.00	0.125	1	U
1,1,1,2-Tetrachloroethane	0.250	1.00	0.250	1	U
1,1,2,2-Tetrachloroethane	0.200	1.00	0.200	1	U
Tetrachloroethene	0.250	1.00	0.250	1	U
Toluene	0.250	1.00	0.250	1	U
1,2,3-Trichlorobenzene	0.150	1.00	0.150	1	U
1,2,4-Trichlorobenzene	0.200	1.00	0.200	1	U
1,1,1-Trichloroethane	0.250	1.00	0.250	1	U
1,1,2-Trichloroethane	0.250	1.00	0.250	1	U
Trichloroethene	0.250	1.00	0.250	1	U
Trichlorofluoromethane	0.250	1.00	0.250	1	U
1,2,3-Trichloropropane	0.500	2.00	0.500	1	U
1,2,4-Trimethylbenzene	0.250	1.00	0.250	1	U
1,3,5-Trimethylbenzene	0.250	1.00	0.250	1	U
Vinyl chloride	0.250	1.00	0.250	1	U
o-Xylene	0.250	1.00	0.250	1	U
m-,p-Xylene	0.500	2.00	0.500	1	U

Surrogates	% Recovery	Surrogate Limits	Qualifier
Dibromofluoromethane	113	85 - 115	PASS
1,2-Dichloroethane-d4	115	70 - 120	PASS
Toluene-d8	102	85 - 120	PASS
4-Bromofluorobenzene	102	75 - 120	PASS

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4831244
30-JUN-2016 15:02

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060927 | Run Date: 06/22/2016 | Sample ID: WG573771-02
Instrument ID: HPMS8 | Run Time: 17:13 | Prep Method: 5030B/5030C/503
File ID: 8M413166 | Analyst: FJB | Method: 8260B
Workgroup (AAB#): WG573771 | Matrix: Water | Units: ug/L
QC Key: DOD4 | Lot#: STD76744 | Cal ID: HPMS8-13-MAY-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Acetone	20.0	22.3	111	40 - 140	
Benzene	20.0	22.4	112	80 - 120	
Bromobenzene	20.0	20.3	102	75 - 125	
Bromochloromethane	20.0	21.5	108	65 - 130	
Bromodichloromethane	20.0	23.6	118	75 - 120	
Bromoform	20.0	19.6	97.8	70 - 130	
Bromomethane	20.0	18.1	90.3	30 - 145	
2-Butanone	20.0	20.3	101	30 - 150	
n-Butylbenzene	20.0	19.7	98.5	70 - 135	
sec-Butylbenzene	20.0	20.9	104	70 - 125	
tert-Butylbenzene	20.0	20.4	102	70 - 130	
Carbon disulfide	20.0	21.9	109	35 - 160	
Carbon tetrachloride	20.0	24.5	123	65 - 140	
Chlorobenzene	20.0	20.6	103	80 - 120	
Chlorodibromomethane	20.0	20.4	102	60 - 135	
Chloroethane	20.0	24.5	122	60 - 135	
Chloroform	20.0	22.8	114	65 - 135	
Chloromethane	20.0	23.9	120	40 - 125	
2-Chlorotoluene	20.0	20.5	102	75 - 125	
4-Chlorotoluene	20.0	20.6	103	75 - 130	
1,2-Dibromo-3-chloropropane	20.0	17.9	89.5	50 - 130	
1,2-Dibromoethane	20.0	20.0	100	80 - 120	
Dibromomethane	20.0	20.8	104	75 - 125	
1,2-Dichlorobenzene	20.0	20.6	103	70 - 120	
1,3-Dichlorobenzene	20.0	20.6	103	75 - 125	
1,4-Dichlorobenzene	20.0	20.5	102	75 - 125	
Dichlorodifluoromethane	20.0	30.6	153	30 - 155	
1,1-Dichloroethane	20.0	21.8	109	70 - 135	
1,2-Dichloroethane	20.0	24.7	123	70 - 130	
1,1-Dichloroethene	20.0	22.6	113	70 - 130	
cis-1,2-Dichloroethene	20.0	22.5	112	70 - 125	
trans-1,2-Dichloroethene	20.0	22.4	112	60 - 140	
1,2-Dichloropropane	20.0	21.7	109	75 - 125	
1,3-Dichloropropane	20.0	21.5	107	75 - 125	
2,2-Dichloropropane	20.0	26.5	133	70 - 135	
cis-1,3-Dichloropropene	20.0	24.8	124	70 - 130	
trans-1,3-Dichloropropene	20.0	21.6	108	55 - 140	
1,1-Dichloropropene	20.0	22.6	113	75 - 130	
Ethylbenzene	20.0	20.3	102	75 - 125	
2-Hexanone	20.0	16.8	84.1	55 - 130	
Hexachlorobutadiene	20.0	21.3	107	50 - 140	

LCS - Modified 03/06/2008
PDF File ID: 4831245
Report generated: 06/30/2016 15:02

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060927 Run Date: 06/22/2016 Sample ID: WG573771-02
Instrument ID: HPMS8 Run Time: 17:13 Prep Method: 5030B/5030C/503
File ID: 8M413166 Analyst: FJB Method: 8260B
Workgroup (AAB#): WG573771 Matrix: Water Units: ug/L
QC Key: DOD4 Lot#: STD76744 Cal ID: HPMS8-13-MAY-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Isopropylbenzene	20.0	21.9	110	75 - 125	
p-Isopropyltoluene	20.0	20.6	103	75 - 130	
4-Methyl-2-pentanone	20.0	20.5	103	60 - 135	
Methylene chloride	20.0	21.6	108	55 - 140	
Naphthalene	20.0	16.4	81.9	55 - 140	
n-Propylbenzene	20.0	21.2	106	70 - 130	
Styrene	20.0	21.2	106	65 - 135	
1,1,1,2-Tetrachloroethane	20.0	21.0	105	80 - 130	
1,1,2,2-Tetrachloroethane	20.0	18.2	91.2	65 - 130	
Tetrachloroethene	20.0	21.5	108	45 - 150	
Toluene	20.0	20.6	103	75 - 120	
1,2,3-Trichlorobenzene	20.0	17.8	89.1	55 - 140	
1,2,4-Trichlorobenzene	20.0	19.3	96.3	65 - 135	
1,1,1-Trichloroethane	20.0	24.6	123	65 - 130	
1,1,2-Trichloroethane	20.0	19.9	99.5	75 - 125	
Trichlorofluoromethane	20.0	25.6	128	60 - 145	
1,2,3-Trichloropropane	20.0	19.5	97.6	75 - 125	
1,2,4-Trimethylbenzene	20.0	20.4	102	75 - 130	
1,3,5-Trimethylbenzene	20.0	21.3	106	75 - 130	
Vinyl chloride	20.0	28.6	143	50 - 145	
o-Xylene	20.0	21.0	105	80 - 120	
m-,p-Xylene	40.0	42.3	106	75 - 130	

Surrogates	% Recovery	Surrogate Limits	Qualifier
Dibromofluoromethane	110	85 - 115	PASS
1,2-Dichloroethane-d4	113	70 - 120	PASS
Toluene-d8	103	85 - 120	PASS
4-Bromofluorobenzene	103	75 - 120	PASS

* EXCEEDS %REC LIMIT

LCS - Modified 03/06/2008
PDF File ID: 4831245
Report generated: 06/30/2016 15:02

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060927 | Run Date: 06/23/2016 | Sample ID: WG573991-02
Instrument ID: HPMS8 | Run Time: 22:28 | Prep Method: 5030B/5030C/503
File ID: 8M413195 | Analyst: FJB | Method: 8260B
Workgroup (AAB#): WG573991 | Matrix: Water | Units: ug/L
QC Key: DOD4 | Lot#: STD76744 | Cal ID: HPMS8-13-MAY-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Acetone	20.0	24.0	120	40 - 140	
Benzene	20.0	21.8	109	80 - 120	
Bromobenzene	20.0	19.4	97.1	75 - 125	
Bromochloromethane	20.0	21.6	108	65 - 130	
Bromodichloromethane	20.0	23.4	117	75 - 120	
Bromoform	20.0	19.1	95.6	70 - 130	
Bromomethane	20.0	17.2	86.1	30 - 145	
2-Butanone	20.0	21.2	106	30 - 150	
n-Butylbenzene	20.0	19.0	94.8	70 - 135	
sec-Butylbenzene	20.0	20.2	101	70 - 125	
tert-Butylbenzene	20.0	20.0	99.8	70 - 130	
Carbon disulfide	20.0	20.9	104	35 - 160	
Carbon tetrachloride	20.0	24.4	122	65 - 140	
Chlorobenzene	20.0	20.7	104	80 - 120	
Chlorodibromomethane	20.0	20.4	102	60 - 135	
Chloroethane	20.0	23.4	117	60 - 135	
Chloroform	20.0	22.7	114	65 - 135	
Chloromethane	20.0	23.0	115	40 - 125	
2-Chlorotoluene	20.0	18.4	92.0	75 - 125	
4-Chlorotoluene	20.0	22.2	111	75 - 130	
1,2-Dibromo-3-chloropropane	20.0	17.1	85.6	50 - 130	
1,2-Dibromoethane	20.0	19.8	98.9	80 - 120	
Dibromomethane	20.0	20.7	104	75 - 125	
1,2-Dichlorobenzene	20.0	19.4	96.9	70 - 120	
1,3-Dichlorobenzene	20.0	20.2	101	75 - 125	
1,4-Dichlorobenzene	20.0	19.8	99.0	75 - 125	
Dichlorodifluoromethane	20.0	28.6	143	30 - 155	
1,1-Dichloroethane	20.0	21.4	107	70 - 135	
1,2-Dichloroethane	20.0	24.7	123	70 - 130	
1,1-Dichloroethene	20.0	22.4	112	70 - 130	
cis-1,2-Dichloroethene	20.0	22.0	110	70 - 125	
trans-1,2-Dichloroethene	20.0	22.1	111	60 - 140	
1,2-Dichloropropane	20.0	21.6	108	75 - 125	
1,3-Dichloropropane	20.0	21.4	107	75 - 125	
2,2-Dichloropropane	20.0	25.6	128	70 - 135	
cis-1,3-Dichloropropene	20.0	24.5	122	70 - 130	
trans-1,3-Dichloropropene	20.0	21.2	106	55 - 140	
1,1-Dichloropropene	20.0	21.8	109	75 - 130	
Ethylbenzene	20.0	20.1	100	75 - 125	
2-Hexanone	20.0	17.3	86.5	55 - 130	
Hexachlorobutadiene	20.0	20.7	103	50 - 140	

LCS - Modified 03/06/2008
PDF File ID: 4831245
Report generated: 06/30/2016 15:02

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060927 Run Date: 06/23/2016 Sample ID: WG573991-02
Instrument ID: HPMS8 Run Time: 22:28 Prep Method: 5030B/5030C/503
File ID: 8M413195 Analyst: FJB Method: 8260B
Workgroup (AAB#): WG573991 Matrix: Water Units: ug/L
QC Key: DOD4 Lot#: STD76744 Cal ID: HPMS8-13-MAY-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Isopropylbenzene	20.0	21.7	109	75 - 125	
p-Isopropyltoluene	20.0	19.8	98.9	75 - 130	
4-Methyl-2-pentanone	20.0	19.9	99.3	60 - 135	
Methylene chloride	20.0	21.3	107	55 - 140	
Naphthalene	20.0	15.2	76.1	55 - 140	
n-Propylbenzene	20.0	20.9	104	70 - 130	
Styrene	20.0	21.1	105	65 - 135	
1,1,1,2-Tetrachloroethane	20.0	20.9	105	80 - 130	
1,1,2,2-Tetrachloroethane	20.0	17.1	85.3	65 - 130	
Tetrachloroethene	20.0	20.9	105	45 - 150	
Toluene	20.0	20.6	103	75 - 120	
1,2,3-Trichlorobenzene	20.0	17.5	87.3	55 - 140	
1,2,4-Trichlorobenzene	20.0	18.5	92.5	65 - 135	
1,1,1-Trichloroethane	20.0	24.7	123	65 - 130	
1,1,2-Trichloroethane	20.0	19.2	95.9	75 - 125	
Trichloroethene	20.0	23.4	117	70 - 125	
Trichlorofluoromethane	20.0	24.5	123	60 - 145	
1,2,3-Trichloropropane	20.0	19.1	95.6	75 - 125	
1,2,4-Trimethylbenzene	20.0	19.9	99.5	75 - 130	
1,3,5-Trimethylbenzene	20.0	20.8	104	75 - 130	
Vinyl chloride	20.0	27.9	139	50 - 145	
o-Xylene	20.0	21.2	106	80 - 120	
m-,p-Xylene	40.0	42.0	105	75 - 130	

Surrogates	% Recovery	Surrogate Limits	Qualifier
Dibromofluoromethane	110	85 - 115	PASS
1,2-Dichloroethane-d4	113	70 - 120	PASS
Toluene-d8	103	85 - 120	PASS
4-Bromofluorobenzene	98.9	75 - 120	PASS

* EXCEEDS %REC LIMIT

LCS - Modified 03/06/2008
PDF File ID: 4831245
Report generated: 06/30/2016 15:02

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060927 Analyst: FJB Prep Method: 5030B/5030C/503
Instrument ID: HPMS8 Matrix: Water Method: 8260B
Workgroup (AAB#): WG573645 Units: ug/L
QC Key: DOD4 Lot #: STD76744

Sample ID: WG573645-02 LCS File ID: 8M413129 Run Date: 06/21/2016 19:50
Sample ID: WG573645-03 LCS2 File ID: 8M413130 Run Date: 06/21/2016 20:19

Analytes	LCS			LCS2			%RPD	%Rec Limits	RPD Lmt	Q
	Known	Found	% REC	Known	Found	% REC				
1,1,1,2-Tetrachloroethane	20.0	21.0	105	20.0	20.4	102	2.55	80 - 130	30	
1,1,1-Trichloroethane	20.0	23.5	118	20.0	23.3	117	0.919	65 - 130	30	
1,1,2,2-Tetrachloroethane	20.0	17.9	89.4	20.0	17.4	87.0	2.68	65 - 130	30	
1,1,2-Trichloroethane	20.0	20.2	101	20.0	19.3	96.3	4.89	75 - 125	30	
1,1-Dichloroethane	20.0	21.9	109	20.0	21.1	106	3.48	70 - 135	30	
1,1-Dichloroethene	20.0	22.3	111	20.0	22.0	110	1.36	70 - 130	30	
1,1-Dichloropropene	20.0	22.7	114	20.0	22.4	112	1.58	75 - 130	30	
1,2,3-Trichlorobenzene	20.0	18.7	93.4	20.0	18.5	92.4	0.993	55 - 140	30	
1,2,3-Trichloropropane	20.0	19.2	95.9	20.0	19.3	96.5	0.580	75 - 125	30	
1,2,4-Trichlorobenzene	20.0	19.7	98.4	20.0	19.2	95.8	2.71	65 - 135	30	
1,2,4-Trimethylbenzene	20.0	20.1	100	20.0	19.8	99.0	1.25	75 - 130	30	
1,2-Dibromo-3-chloropropane	20.0	17.8	88.8	20.0	17.7	88.5	0.377	50 - 130	30	
1,2-Dibromoethane	20.0	20.3	102	20.0	19.7	98.5	3.03	80 - 120	30	
1,2-Dichlorobenzene	20.0	19.9	99.6	20.0	19.5	97.3	2.40	70 - 120	30	
1,2-Dichloroethane	20.0	23.9	119	20.0	23.0	115	3.72	70 - 130	30	
1,2-Dichloropropane	20.0	22.0	110	20.0	21.9	109	0.534	75 - 125	30	
1,3,5-Trimethylbenzene	20.0	21.1	106	20.0	20.8	104	1.31	75 - 130	30	
1,3-Dichlorobenzene	20.0	20.4	102	20.0	20.3	102	0.603	75 - 125	30	
1,3-Dichloropropane	20.0	21.5	108	20.0	20.9	104	3.09	75 - 125	30	
1,4-Dichlorobenzene	20.0	20.4	102	20.0	20.2	101	0.886	75 - 125	30	
2,2-Dichloropropane	20.0	25.5	128	20.0	24.8	124	2.80	70 - 135	30	
2-Butanone	20.0	20.2	101	20.0	19.7	98.5	2.44	30 - 150	30	
2-Chlorotoluene	20.0	20.9	105	20.0	20.7	104	1.14	75 - 125	30	
2-Hexanone	20.0	17.1	85.6	20.0	18.2	90.8	5.87	55 - 130	30	
4-Chlorotoluene	20.0	20.6	103	20.0	19.9	99.7	3.22	75 - 130	30	
4-Methyl-2-pentanone	20.0	21.5	107	20.0	19.0	94.9	12.3	60 - 135	30	
Acetone	20.0	20.1	101	20.0	21.6	108	7.09	40 - 140	30	
Benzene	20.0	22.6	113	20.0	22.1	111	2.24	80 - 120	30	
Bromobenzene	20.0	20.0	99.9	20.0	19.8	99.0	0.968	75 - 125	30	
Bromochloromethane	20.0	22.3	112	20.0	21.3	107	4.47	65 - 130	30	
Bromodichloromethane	20.0	22.9	114	20.0	22.3	111	2.63	75 - 120	30	
Bromoform	20.0	20.1	100	20.0	19.2	96.0	4.39	70 - 130	30	
Bromomethane	20.0	19.0	94.8	20.0	18.4	92.1	2.85	30 - 145	30	
Carbon disulfide	20.0	21.8	109	20.0	21.2	106	3.01	35 - 160	30	
Carbon tetrachloride	20.0	23.8	119	20.0	23.0	115	3.50	65 - 140	30	
Chlorobenzene	20.0	21.3	107	20.0	20.5	103	3.62	80 - 120	30	
Chloroethane	20.0	24.4	122	20.0	23.6	118	3.50	60 - 135	30	
Chloroform	20.0	22.6	113	20.0	21.9	110	2.82	65 - 135	30	
Chloromethane	20.0	23.8	119	20.0	23.4	117	2.06	40 - 125	30	
cis-1,2-Dichloroethene	20.0	22.6	113	20.0	21.8	109	3.37	70 - 125	30	

LCS_LCS2 - Modified 03/06/2008
PDF File ID: 4833867
Report generated: 06/30/2016 15:02

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060927 Analyst: FJB Prep Method: 5030B/5030C/503
Instrument ID: HPMS8 Matrix: Water Method: 8260B
Workgroup (AAB#): WG573645 Units: ug/L
QC Key: DOD4 Lot #: STD76744

Sample ID: WG573645-02 LCS File ID: 8M413129 Run Date: 06/21/2016 19:50
Sample ID: WG573645-03 LCS2 File ID: 8M413130 Run Date: 06/21/2016 20:19

Analytes	LCS			LCS2			%RPD	%Rec Limits	RPD Lmt	Q
	Known	Found	% REC	Known	Found	% REC				
cis-1,3-Dichloropropene	20.0	24.5	123	20.0	24.0	120	2.38	70 - 130	30	
Chlorodibromomethane	20.0	20.6	103	20.0	20.0	100	2.71	60 - 135	30	
Dibromomethane	20.0	20.1	100	20.0	19.9	99.3	1.03	75 - 125	30	
Dichlorodifluoromethane	20.0	28.3	142	20.0	28.0	140	1.27	30 - 155	30	
Ethylbenzene	20.0	20.0	100	20.0	19.8	99.0	1.03	75 - 125	30	
Hexachlorobutadiene	20.0	21.3	106	20.0	21.1	106	0.619	50 - 140	30	
Isopropylbenzene	20.0	22.1	111	20.0	21.9	109	1.21	75 - 125	30	
m-,p-Xylene	40.0	43.1	108	40.0	41.7	104	3.24	75 - 130	30	
Methylene chloride	20.0	21.8	109	20.0	21.3	107	2.40	55 - 140	30	
n-Butylbenzene	20.0	19.8	99.0	20.0	19.6	97.8	1.17	70 - 135	30	
n-Propylbenzene	20.0	21.2	106	20.0	20.9	105	1.09	70 - 130	30	
Naphthalene	20.0	16.1	80.4	20.0	16.5	82.3	2.32	55 - 140	30	
o-Xylene	20.0	21.4	107	20.0	21.1	106	1.42	80 - 120	30	
p-Isopropyltoluene	20.0	20.7	103	20.0	20.1	101	2.77	75 - 130	30	
sec-Butylbenzene	20.0	20.8	104	20.0	20.5	103	1.09	70 - 125	30	
Styrene	20.0	21.3	106	20.0	20.9	104	1.89	65 - 135	30	
tert-Butylbenzene	20.0	21.0	105	20.0	20.4	102	3.11	70 - 130	30	
Tetrachloroethene	20.0	21.6	108	20.0	21.5	108	0.381	45 - 150	30	
Toluene	20.0	21.0	105	20.0	20.7	104	1.65	75 - 120	30	
trans-1,2-Dichloroethene	20.0	22.7	113	20.0	22.2	111	1.86	60 - 140	30	
trans-1,3-Dichloropropene	20.0	21.5	107	20.0	20.5	103	4.66	55 - 140	30	
Trichloroethene	20.0	23.5	118	20.0	23.2	116	1.28	70 - 125	30	
Trichlorofluoromethane	20.0	24.5	122	20.0	23.4	117	4.35	60 - 145	30	
Vinyl chloride	20.0	27.3	137	20.0	27.0	135	1.13	50 - 145	30	

Surogates	LCS	LCS2	Surrogate Limits	Qualifier
	% Recovery	% Recovery		
1,2-Dichloroethane-d4	109	106	70 - 120	PASS
Dibromofluoromethane	107	108	85 - 115	PASS
4-Bromofluorobenzene	98.0	99.0	75 - 120	PASS
Toluene-d8	103	102	85 - 120	PASS

* EXCEEDS %REC LIMIT

\# EXCEEDS RPD LIMIT

LCS_LCS2 - Modified 03/06/2008
PDF File ID: 4833867
Report generated: 06/30/2016 15:02

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

BFB

Login Number: L16060927
Instrument: HPMS8
Analyst: TMB
Workgroup: WG568561

Tune ID: WG568561-01
Run Date: 05/13/2016
Run Time: 12:36
File ID: 8M412218
Cal ID: HPMS8 -

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50.0	95.0	15.0	40.0	16.7	6575	PASS
75.0	95.0	30.0	60.0	51.3	20171	PASS
95.0	95.0	100	100	100	39312	PASS
96.0	95.0	5.00	9.00	6.74	2649	PASS
173	174	0	2.00	0.363	122	PASS
174	95.0	50.0	100	85.5	33616	PASS
175	174	5.00	9.00	8.17	2746	PASS
176	174	95.0	101	99.2	33360	PASS
177	176	5.00	9.00	6.80	2267	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG568561-02	STD	01	05/13/2016 13:29	
WG568561-03	STD	01	05/13/2016 13:58	
WG568561-04	STD	01	05/13/2016 14:27	
WG568561-05	STD	01	05/13/2016 14:56	
WG568561-06	STD	01	05/13/2016 15:25	
WG568561-07	STD	01	05/13/2016 15:54	
WG568561-08	STD-CCV	01	05/13/2016 16:23	
WG568561-09	STD	01	05/13/2016 16:52	
WG568561-10	STD	01	05/13/2016 17:21	
WG568561-11	STD	01	05/13/2016 17:50	
WG568561-12	SSCV	01	05/13/2016 19:17	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4831247
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

BFB

Login Number: L16060927
Instrument: HPMS8
Analyst: FJB
Workgroup: WG573644

Tune ID: WG573644-01
Run Date: 06/21/2016
Run Time: 17:23
File ID: 8M413124
Cal ID: HPMS8 - 13-MAY-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50.0	95.0	15.0	40.0	17.2	10071	PASS
75.0	95.0	30.0	60.0	53.8	31493	PASS
95.0	95.0	100	100	100	58490	PASS
96.0	95.0	5.00	9.00	6.51	3806	PASS
173	174	0	2.00	0.364	174	PASS
174	95.0	50.0	100	81.7	47778	PASS
175	174	5.00	9.00	7.65	3655	PASS
176	174	95.0	101	96.2	45946	PASS
177	176	5.00	9.00	6.54	3004	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573644-02	CCV	01	06/21/2016 18:23	
WG573787-01	CCV	01	06/21/2016 18:52	
WG573645-01	BLANK	01	06/21/2016 19:21	
WG573645-02	LCS	01	06/21/2016 19:50	
WG573645-03	LCS2	01	06/21/2016 20:19	
L16060927-05	TRIP BLANK	01	06/21/2016 22:44	
L16060927-04	18CPTMW19-061516	01	06/22/2016 03:34	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4831247
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

BFB

Login Number: L16060927
Instrument: HPMS8
Analyst: FJB
Workgroup: WG573770

Tune ID: WG573770-01
Run Date: 06/22/2016
Run Time: 14:38
File ID: 8M413161
Cal ID: HPMS8 - 13-MAY-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50.0	95.0	15.0	40.0	19.6	8091	PASS
75.0	95.0	30.0	60.0	56.5	23343	PASS
95.0	95.0	100	100	100	41328	PASS
96.0	95.0	5.00	9.00	6.39	2639	PASS
173	174	0	2.00	0	0	PASS
174	95.0	50.0	100	77.4	31974	PASS
175	174	5.00	9.00	8.17	2612	PASS
176	174	95.0	101	95.1	30422	PASS
177	176	5.00	9.00	7.09	2156	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573770-02	CCV	01	06/22/2016 15:41	
WG573771-01	BLANK	01	06/22/2016 16:39	
WG573771-02	LCS	01	06/22/2016 17:13	
L16060927-02	18CPTMW23-061516	01	06/22/2016 18:40	
WG573023-01	FBLK1	DL01	06/23/2016 03:50	*

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4831247
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

BFB

Login Number: L16060927
Instrument: HPMS8
Analyst: FJB
Workgroup: WG573990

Tune ID: WG573990-01
Run Date: 06/23/2016
Run Time: 19:56
File ID: 8M413190
Cal ID: HPMS8 - 13-MAY-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50.0	95.0	15.0	40.0	19.0	10373	PASS
75.0	95.0	30.0	60.0	58.1	31797	PASS
95.0	95.0	100	100	100	54685	PASS
96.0	95.0	5.00	9.00	6.95	3803	PASS
173	174	0	2.00	0	0	PASS
174	95.0	50.0	100	76.1	41629	PASS
175	174	5.00	9.00	7.18	2991	PASS
176	174	95.0	101	95.9	39941	PASS
177	176	5.00	9.00	6.76	2702	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573990-02	CCV	01	06/23/2016 21:00	
WG573991-01	BLANK	01	06/23/2016 21:59	
WG573991-02	LCS	01	06/23/2016 22:28	
L16060927-02	18CPTMW23-061516	DL01	06/23/2016 22:56	
L16060927-01	18CPTMW23SW-061516	01	06/24/2016 00:24	
L16060927-03	18CPTMW19SW-061516	01	06/24/2016 00:53	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4831247
Report generated 06/30/2016 15:02

Calibration Table Report
Method: 8260WT.M
Title: Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Calibration: Sat May 14 18:08:06 2016
Curve: WG568561
Calibration Files

	Compound	0.3 8M412220.D	0.4 8M412221.D	1 8M412222.D	2 8M412223.D	5 8M412224.D	20 8M412225.D	50 8M412226.D	100 8M412227.D	200 8M412228.D	300 8M412229.D	Avg	%RSD	Linear	Quad
I	Fluorobenzene	ISTD													
T	Dichlorodifluoromethane		0.427	0.390	0.397	0.364	0.468	0.439	0.452	0.427	0.408	0.419	7.751		
P	Chloromethane			0.264	0.257	0.223	0.253	0.240	0.240	0.239	0.241	0.245	5.336		
C	Vinyl Chloride		0.274	0.296	0.297	0.277	0.299	0.292	0.299	0.292	0.288	0.290	3.191		
T	1,3-Butadiene				0.218	0.208	0.242	0.251	0.277	0.254	0.254	0.243	9.644		
T	Bromomethane			0.129	0.133	0.127	0.162	0.167	0.187	0.205	0.212	0.165	20.449		1.000
T	Chloroethane		0.173	0.157	0.153	0.143	0.164	0.157	0.163	0.164	0.164	0.160	5.354		
T	Trichlorofluoromethane		0.544	0.538	0.559	0.506	0.570	0.538	0.551	0.531	0.525	0.540	3.520		
T	Diethyl ether			0.130	0.138	0.123	0.132	0.125	0.128		0.127	0.129	3.866		
T	Isoprene		0.335	0.351	0.334	0.293	0.334	0.320	0.334	0.322	0.329	0.328	4.867		
T	Acrolein				0.012	0.010	0.010	0.011	0.012		0.014	0.012	12.987		
T	1,1,2-Trichloro-1,2,2-Trifluoroethane			0.256	0.269	0.240	0.260	0.249	0.257	0.251	0.250	0.254	3.397		
T	Acetone					0.023	0.030	0.029	0.030	0.028	0.028	0.028	9.132		
C	1,1-Dichloroethene		0.424	0.407	0.406	0.387	0.416	0.398	0.409	0.400	0.393	0.404	2.830		
T	Tert-Butyl Alcohol				0.010	0.009	0.009	0.009	0.009		0.009	0.009	2.872		
T	Dimethyl Sulfide		0.133	0.142	0.152	0.137	0.145	0.142	0.151	0.147	0.151	0.145	4.534		
T	Iodomethane				0.019	0.024	0.066	0.101	0.129	0.135	0.131	0.087	58.244		0.998
T	Methyl acetate					0.075	0.076	0.074	0.078	0.076	0.078	0.076	2.221		
T	Methylene Chloride			0.274	0.265	0.244	0.263	0.250	0.255	0.254	0.252	0.257	3.686		
T	Carbon Disulfide			0.888	0.922	0.801	0.873	0.816	0.854	0.813	0.802	0.846	5.346		
T	Acrylonitrile			0.027	0.035	0.033	0.038	0.038	0.039		0.044	0.036	14.671		
T	Methyl Tert Butyl Ether			0.534	0.543	0.521	0.586	0.579	0.585	0.560	0.546	0.557	4.455		
T	trans-1,2-Dichloroethene		0.389	0.395	0.364	0.352	0.381	0.367	0.372	0.364	0.358	0.371	3.858		
T	n-Hexane					0.265	0.280	0.270	0.286	0.275	0.279	0.276	2.732		
T	Diisopropyl ether			0.600	0.607	0.548	0.588	0.572	0.571		0.567	0.579	3.559		
T	Vinyl Acetate				0.256	0.243	0.270	0.275	0.266	0.271	0.275	0.265	4.509		
P	1,1-Dichloroethane		0.441	0.463	0.486	0.434	0.488	0.464	0.470	0.459	0.447	0.461	4.012		
T	Ethyl-Tert-Butyl ether			0.616	0.645	0.590	0.641	0.628	0.619		0.603	0.620	3.188		
T	2-Butanone					0.035	0.042	0.043	0.045	0.041	0.043	0.041	8.268		
T	Propionitrile			0.010	0.012	0.012	0.013	0.012	0.013		0.013	0.012	8.354		
T	2,2-Dichloropropane		0.471	0.462	0.520	0.476	0.503	0.494	0.497	0.483	0.470	0.486	3.839		
T	cis-1,2-Dichloroethene		0.299	0.298	0.291	0.267	0.295	0.282	0.289	0.283	0.277	0.287	3.685		
C	Chloroform	0.563	0.567	0.581	0.536	0.509	0.566	0.543	0.546	0.529	0.506	0.545	4.619		
T	1-Bromopropane			0.034	0.040	0.040	0.043	0.041	0.043	0.043	0.044	0.041	7.649		
T	Bromochloromethane		0.131	0.164	0.147	0.137	0.154	0.148	0.151	0.148	0.145	0.147	6.598		
T	Tetrahydrofuran			0.034	0.028	0.025	0.026	0.026	0.026		0.026	0.027	11.724		
S	Dibromofluoromethane			0.249	0.261	0.257	0.280	0.271	0.284	0.271	0.275	0.269	4.465		
T	1,1,1-Trichloroethane		0.492	0.537	0.515	0.496	0.545	0.525	0.533	0.515	0.499	0.517	3.660		
T	Cyclohexane			0.373	0.369	0.335	0.359	0.344	0.366	0.356	0.364	0.358	3.619		
T	1,1-Dichloropropene			0.414	0.399	0.369	0.403	0.395	0.399	0.392	0.382	0.394	3.488		
T	Tert-Amyl-Methyl ether			0.567	0.587	0.534	0.576	0.568	0.561		0.549	0.563	3.077		
T	Carbon Tetrachloride		0.444	0.456	0.462	0.441	0.481	0.465	0.474	0.462	0.448	0.4594	2.92173		
S	1,2-Dichloroethane-d4			0.241	0.309	0.276	0.312	0.295	0.301	0.284	0.286	0.288	7.93644		
T	Heptane											0	0		
T	1,2-Dichloroethane		0.347	0.335	0.354	0.336	0.371	0.36	0.364	0.349	0.338	0.3505	3.65503		
T	Benzene		1.128	1.084	1.071	0.977	1.085	1.033	1.019	0.963	0.883	1.0269	7.39343		
T	Trichloroethene		0.285	0.275	0.276	0.251	0.283	0.273	0.279	0.273	0.27	0.2739	3.66126		
T	Methylcyclohexane					0.396	0.426	0.408	0.429	0.416	0.422	0.4161	2.94153		
C	1,2-Dichloropropane		0.239	0.222	0.233	0.207	0.238	0.226	0.233	0.231	0.229	0.2286	4.30096		
T	Bromodichloromethane		0.369	0.382	0.412	0.385	0.426	0.421	0.417	0.41	0.395	0.4018	4.93305		
T	1,4-Dioxane					0.001	0.001	0.001	0.001		0.001	0.0011	14.5817		
T	Dibromomethane		0.141	0.13	0.137	0.128	0.141	0.134	0.137	0.136	0.133	0.1353	3.26799		
T	2-Chloroethyl Vinyl Ether				0.1	0.084	0.103	0.104	0.103	0.095	0.098	0.0983	7.09792		
T	4-Methyl-2-Pentanone					0.033	0.042	0.042	0.043	0.04	0.041	0.0401	8.66859		
T	cis-1,3-Dichloropropene		0.391	0.393	0.386	0.398	0.431	0.428	0.428	0.42	0.405	0.409	4.41527		

T	Dimethyl Disulfide					0.182	0.218	0.221	0.23	0.229	0.23	0.2183	8.49419	
I	Chlorobenzene-d5	ISTD												
S	Toluene-d8			1.282	1.295	1.228	1.385	1.333	1.378	1.289	1.278	1.3084	4.07746	
C	Toluene		1.653	1.696	1.678	1.507	1.647	1.554	1.545	1.418	1.27	1.552	8.98325	
T	Ethyl Methacrylate			0.251	0.293	0.268	0.314	0.314	0.327	0.309	0.31	0.2981	8.76897	
T	Paraldehyde											0	0	
T	trans-1,3-Dichloropropene			0.472	0.51	0.466	0.561	0.552	0.553	0.533	0.515	0.5201	7.01006	
T	1,1,2-Trichloroethane		0.261	0.241	0.242	0.222	0.254	0.249	0.252	0.243	0.241	0.2449	4.54209	
T	2-Hexanone					0.042	0.053	0.053	0.056	0.051	0.053	0.0515	9.50479	
T	1,3-Dichloropropane		0.355	0.442	0.465	0.42	0.468	0.454	0.457	0.437	0.428	0.436	7.93348	
T	Tetrachloroethene		0.355	0.337	0.336	0.306	0.333	0.325	0.337	0.334	0.33	0.3325	3.83886	
T	Dibromochloromethane		0.349	0.336	0.338	0.328	0.373	0.372	0.383	0.373	0.367	0.3576	5.60043	
T	1,2-Dibromoethane		0.233	0.234	0.234	0.227	0.259	0.254	0.258	0.251	0.247	0.2441	5.04151	
T	1-Chlorohexane		0.532	0.584	0.505	0.502	0.54	0.52	0.547	0.521	0.516	0.5298	4.7826	
P	Chlorobenzene		1.029	1.173	1.09	1.007	1.099	1.037	1.043	1.002	0.939	1.0466	6.41465	
T	1,1,1,2-Tetrachloroethane		0.392	0.399	0.403	0.357	0.431	0.414	0.424	0.417	0.413	0.4055	5.39495	
C	Ethylbenzene		0.6	0.594	0.59	0.532	0.586	0.559	0.57	0.564	0.558	0.5726	3.7769	
T	m-,p-Xylene		0.744	0.696	0.691	0.643	0.713	0.67	0.673	0.634	0.581	0.6716	7.12586	
T	o-Xylene			0.691	0.651	0.614	0.692	0.658	0.667	0.65	0.63	0.6567	4.11758	
T	Styrene		1.05	1.06	1.096	1.043	1.159	1.124	1.12	1.061	0.98	1.077	4.96781	
P	Bromoform			0.177	0.21	0.181	0.226	0.221	0.232	0.227	0.223	0.2121	10.0546	
T	Isopropylbenzene		1.853	1.891	1.871	1.708	1.897	1.781	1.75	1.573	1.377	1.7444	9.92447	
I	1,4-Dichlorobenzene-d4	ISTD												
P	1,1,2,2-Tetrachloroethane		0.426	0.492	0.503	0.486	0.543	0.51	0.527	0.513	0.514	0.5018	6.60291	
S	p-Bromofluorobenzene				1.054	0.976	1.071	1.016	1.053	1.031	1.046	1.0352	3.05136	
T	1,2,3-Trichloropropane			0.119	0.158	0.161	0.162	0.155	0.162	0.16	0.158	0.1545	9.40789	
T	trans-1,4-Dichloro-2-Butene			0.085	0.118	0.12	0.149	0.151	0.16	0.159	0.161	0.1376	19.9543	1.000
T	n-Propylbenzene		4.139	4.239	4.243	4.051	4.241	3.953	3.882	3.419	2.935	3.9003	11.4439	
T	Bromobenzene	0.796	0.708	0.843	0.837	0.778	0.84	0.798	0.817	0.803	0.803	0.8022	4.93253	
T	1,3,5-Trimethylbenzene		2.941	2.93	2.941	2.839	3.053	2.882	2.871	2.689	2.423	2.8411	6.50391	
T	2-Chlorotoluene		2.917	2.935	2.929	2.712	2.861	2.795	2.682	2.533	2.145	2.7231	9.37022	
T	4-Chlorotoluene		2.681	2.596	2.581	2.474	2.625	2.326	2.447	2.286	2.121	2.4597	7.49558	
T	a-Methylstyrene					1.257	1.406	1.34	1.384	1.35	1.351	1.3479	3.77654	
T	tert-Butylbenzene			0.587	0.555	0.53	0.571	0.533	0.552	0.561	0.558	0.5558	3.34508	
T	1,2,4-Trimethylbenzene		2.955	3.168	3.127	2.936	3.12	2.914	2.937	2.709	2.427	2.9214	7.97314	
T	sec-Butylbenzene			3.701	3.536	3.487	3.607	3.34	3.332	3.045	2.69	3.3424	9.93324	
T	p-Isopropyltoluene			3.112	3.062	2.903	2.988	2.813	2.809	2.634	2.382	2.8379	8.45265	
T	1,3-Dichlorobenzene		1.484	1.617	1.627	1.53	1.605	1.531	1.549	1.51	1.45	1.5448	3.95774	
T	1,4-Dichlorobenzene	1.713	1.524	1.672	1.594	1.503	1.6	1.511	1.535	1.49	1.431	1.5574	5.56562	
T	n-Butylbenzene			3.143	2.96	2.895	3.002	2.791	2.792	2.571	2.292	2.8056	9.55812	
T	1,2-Dichlorobenzene	1.59	1.307	1.415	1.41	1.322	1.404	1.333	1.356	1.315	1.286	1.3739	6.4618	
T	1,2-Dibromo-3-Chloropropane				0.085	0.103	0.102	0.103	0.106	0.104	0.103	0.1007	6.86771	
T	1,2,4-Trichlorobenzene		1.074	1.01	1.028	0.982	1.069	1.013	1.036	1.029	1.002	1.0269	2.90829	
T	Hexachlorobutadiene		0.516	0.476	0.483	0.447	0.487	0.459	0.48	0.485	0.485	0.4798	4.00465	
T	Naphthalene		1.594	1.457	1.574	1.455	1.63	1.552	1.564	1.485	1.422	1.5258	4.75891	
T	1,2,3-Trichlorobenzene	0.913	0.934	0.905	0.862	0.805	0.877	0.824	0.862	0.846	0.835	0.8664	4.7474	

Tue May 17 11:49:41 2016

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16060927 Run Date: 05/13/2016 Sample ID: WG568561-12
Instrument ID: HPMS8 Run Time: 19:17 Method: 8260B
File ID: 8M412232 Analyst: TMB QC Key: DOD4
ICal Workgroup: WG568561 Cal ID: HPMS8 - 13-MAY-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
Chloroform	CCC	50.0	47.1	ug/L	0.513	5.70	20	
1,1-Dichloroethene	CCC	50.0	46.0	ug/L	0.372	8.10	20	
1,2-Dichloropropane	CCC	50.0	51.0	ug/L	0.233	1.90	20	
Ethylbenzene	CCC	50.0	48.2	ug/L	0.552	3.60	20	
Toluene	CCC	50.0	49.1	ug/L	1.52	1.80	20	
Vinyl Chloride	CCC	50.0	52.4	ug/L	0.304	4.80	20	
Bromoform	SPCC	50.0	48.2	ug/L	0.205	3.60	20	
Chlorobenzene	SPCC	50.0	50.2	ug/L	1.05	0.400	20	
Chloromethane	SPCC	50.0	50.7	ug/L	0.248	1.40	20	
1,1-Dichloroethane	SPCC	50.0	46.7	ug/L	0.431	6.60	20	
1,1,2,2-Tetrachloroethane	SPCC	50.0	49.6	ug/L	0.498	0.800	20	
Acetone		50.0	57.5	ug/L	0.0324	15.1	20	
Benzene		50.0	48.5	ug/L	0.997	3.00	20	
Bromobenzene		50.0	49.5	ug/L	0.794	1.10	20	
Bromochloromethane		50.0	47.5	ug/L	0.140	4.90	20	
Bromodichloromethane		50.0	48.3	ug/L	0.388	3.40	20	
Bromomethane		50.0	46.4	ug/L	0.162	7.20	20	
2-Butanone		50.0	53.0	ug/L	0.0438	6.00	20	
n-Butylbenzene		50.0	49.7	ug/L	2.79	0.700	20	
sec-Butylbenzene		50.0	51.3	ug/L	3.43	2.70	20	
tert-Butylbenzene		50.0	50.9	ug/L	0.566	1.90	20	
Carbon Disulfide		50.0	42.3	ug/L	0.716	15.4	20	
Carbon Tetrachloride		50.0	46.2	ug/L	0.425	7.50	20	
Dibromochloromethane		50.0	49.7	ug/L	0.355	0.700	20	
Chloroethane		50.0	53.3	ug/L	0.170	6.70	20	
2-Chlorotoluene		50.0	51.6	ug/L	2.81	3.10	20	
4-Chlorotoluene		50.0	48.3	ug/L	2.37	3.50	20	
1,2-Dibromo-3-Chloropropane		50.0	48.5	ug/L	0.0977	3.00	20	
1,2-Dibromoethane		50.0	51.1	ug/L	0.250	2.20	20	
Dibromomethane		50.0	45.8	ug/L	0.124	8.40	20	
1,2-Dichlorobenzene		50.0	50.5	ug/L	1.39	1.00	20	
1,3-Dichlorobenzene		50.0	50.9	ug/L	1.57	1.80	20	
1,4-Dichlorobenzene		50.0	50.5	ug/L	1.57	0.900	20	
Dichlorodifluoromethane		50.0	54.2	ug/L	0.455	8.40	20	
1,2-Dichloroethane		50.0	49.0	ug/L	0.344	2.00	20	
cis-1,2-Dichloroethene		50.0	48.9	ug/L	0.281	2.20	20	
trans-1,2-Dichloroethene		50.0	48.1	ug/L	0.357	3.90	20	
1,3-Dichloropropane		50.0	53.7	ug/L	0.468	7.30	20	
2,2-Dichloropropane		50.0	47.9	ug/L	0.466	4.20	20	
cis-1,3-Dichloropropene		50.0	53.9	ug/L	0.441	7.80	20	
trans-1,3-Dichloropropene		50.0	50.7	ug/L	0.527	1.30	20	
1,1-Dichloropropene		50.0	47.5	ug/L	0.374	5.10	20	

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4831246
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16060927 Run Date: 05/13/2016 Sample ID: WG568561-12
Instrument ID: HPMS8 Run Time: 19:17 Method: 8260B
File ID: 8M412232 Analyst: TMB QC Key: DOD4
ICal Workgroup: WG568561 Cal ID: HPMS8 - 13-MAY-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
2-Hexanone		50.0	53.6	ug/L	0.0552	7.20	20	
Hexachlorobutadiene		50.0	52.1	ug/L	0.500	4.20	20	
Isopropylbenzene		50.0	51.1	ug/L	1.78	2.20	20	
p-Isopropyltoluene		50.0	50.1	ug/L	2.85	0.300	20	
4-Methyl-2-Pentanone		50.0	51.9	ug/L	0.0416	3.70	20	
Methylene Chloride		50.0	48.4	ug/L	0.249	3.10	20	
Naphthalene		50.0	45.8	ug/L	1.40	8.40	20	
n-Propylbenzene		50.0	51.9	ug/L	4.05	3.80	20	
Styrene		50.0	51.5	ug/L	1.11	3.00	20	
1,1,1,2-Tetrachloroethane		50.0	49.4	ug/L	0.400	1.20	20	
Tetrachloroethene		50.0	47.8	ug/L	0.318	4.40	20	
1,2,3-Trichlorobenzene		50.0	47.7	ug/L	0.827	4.50	20	
1,2,4-Trichlorobenzene		50.0	50.1	ug/L	1.03	0.200	20	
1,1,1-Trichloroethane		50.0	47.6	ug/L	0.492	4.80	20	
1,1,2-Trichloroethane		50.0	49.0	ug/L	0.240	1.90	20	
Trichloroethene		50.0	51.1	ug/L	0.280	2.10	20	
Trichlorofluoromethane		50.0	46.9	ug/L	0.507	6.10	20	
1,2,3-Trichloropropane		50.0	50.1	ug/L	0.155	0.200	20	
1,2,4-Trimethylbenzene		50.0	49.7	ug/L	2.90	0.700	20	
1,3,5-Trimethylbenzene		50.0	51.6	ug/L	2.93	3.30	20	
o-Xylene		50.0	50.6	ug/L	0.664	1.10	20	
m-,p-Xylene		100	101	ug/L	0.679	1.10	20	

* Exceeds %D Limit

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4831246
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.
CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927 Run Date: 06/21/2016 Sample ID: WG573644-02
Instrument ID: HPMS8 Run Time: 18:23 Method: 8260B
File ID: 8M413126 Analyst: FJB QC Key: DOD4
Workgroup (AAB#): WG573645 Cal ID: HPMS8 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Chloroform	CCC	50.0	53.3	ug/L	0.580	6.54	20	
1,1-Dichloroethene	CCC	50.0	53.5	ug/L	0.433	7.04	20	
1,2-Dichloropropane	CCC	50.0	50.3	ug/L	0.230	0.500	20	
Ethylbenzene	CCC	50.0	47.5	ug/L	0.544	4.94	20	
Toluene	CCC	50.0	48.5	ug/L	1.51	3.03	20	
Vinyl Chloride	CCC	50.0	54.4	ug/L	0.316	8.72	20	
Bromoform	SPCC	50.0	52.5	ug/L	0.223	5.01	20	
Chlorobenzene	SPCC	50.0	48.2	ug/L	1.01	3.50	20	
Chloromethane	SPCC	50.0	46.8	ug/L	0.229	6.35	20	
1,1-Dichloroethane	SPCC	50.0	53.0	ug/L	0.489	5.95	20	
1,1,2,2-Tetrachloroethane	SPCC	50.0	45.0	ug/L	0.452	10.0	20	
Xylenes		150	146	ug/L	0.644	2.96	20	
Acetone		50.0	57.7	ug/L	0.0325	15.4	20	
Benzene		50.0	51.8	ug/L	1.06	3.56	20	
Bromobenzene		50.0	48.5	ug/L	0.779	2.91	20	
Bromochloromethane		50.0	54.1	ug/L	0.159	8.24	20	
Bromodichloromethane		50.0	54.1	ug/L	0.435	8.14	20	
Bromomethane		50.0	40.5	ug/L	0.141	18.9	20	
2-Butanone		50.0	48.6	ug/L	0.0401	2.87	20	
n-Butylbenzene		50.0	47.3	ug/L	2.65	5.48	20	
sec-Butylbenzene		50.0	47.6	ug/L	3.18	4.85	20	
tert-Butylbenzene		50.0	47.5	ug/L	0.528	5.02	20	
Carbon Disulfide		50.0	53.3	ug/L	0.903	6.70	20	
Carbon Tetrachloride		50.0	56.0	ug/L	0.514	11.9	20	
Dibromochloromethane		50.0	50.2	ug/L	0.359	0.303	20	
Chloroethane		50.0	52.5	ug/L	0.168	5.06	20	
2-Chlorotoluene		50.0	48.3	ug/L	2.63	3.37	20	
4-Chlorotoluene		50.0	46.4	ug/L	2.28	7.20	20	
1,2-Dibromo-3-Chloropropane		50.0	45.5	ug/L	0.0916	9.09	20	
1,2-Dibromoethane		50.0	47.6	ug/L	0.233	4.74	20	
Dibromomethane		50.0	50.8	ug/L	0.137	1.57	20	
1,2-Dichlorobenzene		50.0	46.7	ug/L	1.28	6.54	20	
1,3-Dichlorobenzene		50.0	48.5	ug/L	1.50	2.95	20	
1,4-Dichlorobenzene		50.0	47.3	ug/L	1.47	5.34	20	
Dichlorodifluoromethane		50.0	50.1	ug/L	0.420	0.103	20	
1,2-Dichloroethane		50.0	54.9	ug/L	0.385	9.81	20	
cis-1,2-Dichloroethene		50.0	52.2	ug/L	0.300	4.49	20	
trans-1,2-Dichloroethene		50.0	52.6	ug/L	0.391	5.28	20	
1,3-Dichloropropane		50.0	48.2	ug/L	0.420	3.60	20	
2,2-Dichloropropane		50.0	59.0	ug/L	0.573	18.0	20	
cis-1,3-Dichloropropene		50.0	53.6	ug/L	0.438	7.12	20	
trans-1,3-Dichloropropene		50.0	50.7	ug/L	0.528	1.50	20	

CCV - Modified 03/05/2008
PDF File ID: 4831248
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927 Run Date: 06/21/2016 Sample ID: WG573644-02
Instrument ID: HPMS8 Run Time: 18:23 Method: 8260B
File ID: 8M413126 Analyst: FJB QC Key: DOD4
Workgroup (AAB#): WG573645 Cal ID: HPMS8 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
1,1-Dichloropropene		50.0	53.6	ug/L	0.423	7.28	20	
2-Hexanone		50.0	46.4	ug/L	0.0478	7.10	20	
Hexachlorobutadiene		50.0	48.7	ug/L	0.467	2.62	20	
Isopropylbenzene		50.0	50.1	ug/L	1.75	0.245	20	
p-Isopropyltoluene		50.0	48.3	ug/L	2.74	3.38	20	
4-Methyl-2-Pentanone		50.0	49.6	ug/L	0.0398	0.795	20	
Methylene Chloride		50.0	50.9	ug/L	0.262	1.79	20	
Naphthalene		50.0	45.1	ug/L	1.38	9.72	20	
n-Propylbenzene		50.0	48.3	ug/L	3.77	3.36	20	
Styrene		50.0	49.6	ug/L	1.07	0.785	20	
1,1,1,2-Tetrachloroethane		50.0	49.9	ug/L	0.404	0.290	20	
Tetrachloroethene		50.0	52.0	ug/L	0.346	3.96	20	
1,2,3-Trichlorobenzene		50.0	46.5	ug/L	0.805	7.04	20	
1,2,4-Trichlorobenzene		50.0	49.1	ug/L	1.01	1.80	20	
1,1,1-Trichloroethane		50.0	55.7	ug/L	0.576	11.4	20	
1,1,2-Trichloroethane		50.0	47.4	ug/L	0.232	5.26	20	
Trichloroethene		50.0	53.9	ug/L	0.295	7.70	20	
Trichlorofluoromethane		50.0	55.5	ug/L	0.600	11.0	20	
1,2,3-Trichloropropane		50.0	46.8	ug/L	0.145	6.47	20	
1,2,4-Trimethylbenzene		50.0	48.0	ug/L	2.81	3.94	20	
1,3,5-Trimethylbenzene		50.0	48.7	ug/L	2.77	2.62	20	
o-Xylene		50.0	48.5	ug/L	0.637	3.01	20	
m-,p-Xylene		100	97.1	ug/L	0.652	2.93	20	

* Exceeds %D Criteria

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

CCV - Modified 03/05/2008
PDF File ID: 4831248
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.
CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927 Run Date: 06/22/2016 Sample ID: WG573770-02
Instrument ID: HPMS8 Run Time: 15:41 Method: 8260B
File ID: 8M413163 Analyst: FJB QC Key: DOD4
Workgroup (AAB#): WG573771 Cal ID: HPMS8 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Chloroform	CCC	50.0	55.7	ug/L	0.607	11.5	20	
1,1-Dichloroethene	CCC	50.0	56.9	ug/L	0.460	13.7	20	
1,2-Dichloropropane	CCC	50.0	50.5	ug/L	0.231	0.966	20	
Ethylbenzene	CCC	50.0	49.7	ug/L	0.569	0.554	20	
Toluene	CCC	50.0	50.6	ug/L	1.57	1.21	20	
Vinyl Chloride	CCC	50.0	57.6	ug/L	0.334	15.1	20	
Bromoform	SPCC	50.0	53.4	ug/L	0.226	6.72	20	
Chlorobenzene	SPCC	50.0	50.0	ug/L	1.05	0.0460	20	
Chloromethane	SPCC	50.0	49.7	ug/L	0.243	0.692	20	
1,1-Dichloroethane	SPCC	50.0	55.2	ug/L	0.509	10.4	20	
1,1,2,2-Tetrachloroethane	SPCC	50.0	43.6	ug/L	0.438	12.8	20	
Xylenes		150	153	ug/L	0.679	2.23	20	
Acetone		50.0	58.5	ug/L	0.0329	17.1	20	
Benzene		50.0	53.3	ug/L	1.09	6.55	20	
Bromobenzene		50.0	48.3	ug/L	0.775	3.39	20	
Bromochloromethane		50.0	54.5	ug/L	0.160	8.95	20	
Bromodichloromethane		50.0	57.0	ug/L	0.458	13.9	20	
Bromomethane		50.0	36.7	ug/L	0.127	26.7	20	*
2-Butanone		50.0	48.2	ug/L	0.0399	3.53	20	
n-Butylbenzene		50.0	47.9	ug/L	2.69	4.27	20	
sec-Butylbenzene		50.0	48.0	ug/L	3.21	3.97	20	
tert-Butylbenzene		50.0	47.1	ug/L	0.524	5.78	20	
Carbon Disulfide		50.0	58.5	ug/L	0.990	17.0	20	
Carbon Tetrachloride		50.0	60.9	ug/L	0.560	21.8	20	*
Dibromochloromethane		50.0	52.0	ug/L	0.372	3.99	20	
Chloroethane		50.0	52.8	ug/L	0.169	5.64	20	
2-Chlorotoluene		50.0	49.6	ug/L	2.70	0.747	20	
4-Chlorotoluene		50.0	46.7	ug/L	2.30	6.56	20	
1,2-Dibromo-3-Chloropropane		50.0	45.0	ug/L	0.0907	9.95	20	
1,2-Dibromoethane		50.0	49.4	ug/L	0.241	1.19	20	
Dibromomethane		50.0	50.7	ug/L	0.137	1.43	20	
1,2-Dichlorobenzene		50.0	46.7	ug/L	1.28	6.56	20	
1,3-Dichlorobenzene		50.0	48.7	ug/L	1.50	2.60	20	
1,4-Dichlorobenzene		50.0	48.0	ug/L	1.49	4.08	20	
Dichlorodifluoromethane		50.0	53.9	ug/L	0.452	7.84	20	
1,2-Dichloroethane		50.0	58.1	ug/L	0.407	16.1	20	
cis-1,2-Dichloroethene		50.0	53.4	ug/L	0.306	6.71	20	
trans-1,2-Dichloroethene		50.0	54.3	ug/L	0.404	8.64	20	
1,3-Dichloropropane		50.0	48.3	ug/L	0.421	3.37	20	
2,2-Dichloropropane		50.0	63.7	ug/L	0.619	27.3	20	*
cis-1,3-Dichloropropene		50.0	55.8	ug/L	0.456	11.6	20	
trans-1,3-Dichloropropene		50.0	52.5	ug/L	0.546	5.05	20	

CCV - Modified 03/05/2008
PDF File ID: 4831248
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927 Run Date: 06/22/2016 Sample ID: WG573770-02
Instrument ID: HPMS8 Run Time: 15:41 Method: 8260B
File ID: 8M413163 Analyst: FJB QC Key: DOD4
Workgroup (AAB#): WG573771 Cal ID: HPMS8 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
1,1-Dichloropropene		50.0	56.4	ug/L	0.445	12.8	20	
2-Hexanone		50.0	44.7	ug/L	0.0460	10.6	20	
Hexachlorobutadiene		50.0	48.0	ug/L	0.461	3.92	20	
Isopropylbenzene		50.0	52.2	ug/L	1.82	4.34	20	
p-Isopropyltoluene		50.0	48.7	ug/L	2.76	2.68	20	
4-Methyl-2-Pentanone		50.0	47.2	ug/L	0.0379	5.67	20	
Methylene Chloride		50.0	50.8	ug/L	0.261	1.57	20	
Naphthalene		50.0	43.2	ug/L	1.32	13.5	20	
n-Propylbenzene		50.0	48.6	ug/L	3.79	2.73	20	
Styrene		50.0	51.9	ug/L	1.12	3.79	20	
1,1,1,2-Tetrachloroethane		50.0	52.0	ug/L	0.421	3.93	20	
Tetrachloroethene		50.0	54.4	ug/L	0.362	8.81	20	
1,2,3-Trichlorobenzene		50.0	44.1	ug/L	0.765	11.7	20	
1,2,4-Trichlorobenzene		50.0	47.3	ug/L	0.971	5.49	20	
1,1,1-Trichloroethane		50.0	59.7	ug/L	0.618	19.4	20	
1,1,2-Trichloroethane		50.0	48.0	ug/L	0.235	4.02	20	
Trichlorofluoromethane		50.0	60.6	ug/L	0.655	21.2	20	*
1,2,3-Trichloropropane		50.0	45.8	ug/L	0.142	8.31	20	
1,2,4-Trimethylbenzene		50.0	48.9	ug/L	2.86	2.17	20	
1,3,5-Trimethylbenzene		50.0	49.8	ug/L	2.83	0.374	20	
o-Xylene		50.0	51.2	ug/L	0.673	2.41	20	
m-,p-Xylene		100	102	ug/L	0.686	2.14	20	

* Exceeds %D Criteria

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

CCV - Modified 03/05/2008
PDF File ID: 4831248
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927 Run Date: 06/23/2016 Sample ID: WG573990-02
Instrument ID: HPMS8 Run Time: 21:00 Method: 8260B
File ID: 8M413192 Analyst: FJB QC Key: DOD4
Workgroup (AAB#): WG573991 Cal ID: HPMS8 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Chloroform	CCC	50.0	54.9	ug/L	0.598	9.72	20	
1,1-Dichloroethene	CCC	50.0	54.2	ug/L	0.439	8.49	20	
1,2-Dichloropropane	CCC	50.0	49.4	ug/L	0.226	1.11	20	
Ethylbenzene	CCC	50.0	46.8	ug/L	0.536	6.46	20	
Toluene	CCC	50.0	47.3	ug/L	1.47	5.48	20	
Vinyl Chloride	CCC	50.0	56.0	ug/L	0.325	12.0	20	
Bromoform	SPCC	50.0	50.5	ug/L	0.214	0.904	20	
Chlorobenzene	SPCC	50.0	47.1	ug/L	0.986	5.76	20	
Chloromethane	SPCC	50.0	47.4	ug/L	0.232	5.29	20	
1,1-Dichloroethane	SPCC	50.0	53.1	ug/L	0.490	6.25	20	
1,1,2,2-Tetrachloroethane	SPCC	50.0	43.2	ug/L	0.433	13.7	20	
Xylenes		150	143	ug/L	0.633	4.33	20	
Acetone		50.0	55.7	ug/L	0.0313	11.4	20	
Benzene		50.0	51.7	ug/L	1.06	3.36	20	
Bromobenzene		50.0	48.2	ug/L	0.773	3.63	20	
Bromochloromethane		50.0	53.3	ug/L	0.157	6.63	20	
Bromodichloromethane		50.0	56.3	ug/L	0.452	12.6	20	
Bromomethane		50.0	44.2	ug/L	0.154	11.6	20	
2-Butanone		50.0	47.8	ug/L	0.0395	4.39	20	
n-Butylbenzene		50.0	47.6	ug/L	2.67	4.73	20	
sec-Butylbenzene		50.0	48.0	ug/L	3.21	3.91	20	
tert-Butylbenzene		50.0	46.7	ug/L	0.520	6.51	20	
Carbon Disulfide		50.0	55.6	ug/L	0.940	11.2	20	
Carbon Tetrachloride		50.0	60.0	ug/L	0.551	19.9	20	
Dibromochloromethane		50.0	49.7	ug/L	0.356	0.505	20	
Chloroethane		50.0	49.8	ug/L	0.159	0.424	20	
2-Chlorotoluene		50.0	47.6	ug/L	2.59	4.80	20	
4-Chlorotoluene		50.0	47.6	ug/L	2.34	4.71	20	
1,2-Dibromo-3-Chloropropane		50.0	44.8	ug/L	0.0902	10.5	20	
1,2-Dibromoethane		50.0	47.7	ug/L	0.233	4.68	20	
Dibromomethane		50.0	50.5	ug/L	0.137	0.985	20	
1,2-Dichlorobenzene		50.0	46.5	ug/L	1.28	7.08	20	
1,3-Dichlorobenzene		50.0	48.2	ug/L	1.49	3.65	20	
1,4-Dichlorobenzene		50.0	47.0	ug/L	1.46	5.94	20	
Dichlorodifluoromethane		50.0	52.4	ug/L	0.439	4.74	20	
1,2-Dichloroethane		50.0	58.3	ug/L	0.409	16.6	20	
cis-1,2-Dichloroethene		50.0	52.0	ug/L	0.298	3.98	20	
trans-1,2-Dichloroethene		50.0	52.6	ug/L	0.391	5.23	20	
1,3-Dichloropropane		50.0	46.4	ug/L	0.405	7.22	20	
2,2-Dichloropropane		50.0	61.6	ug/L	0.599	23.3	20	*
cis-1,3-Dichloropropene		50.0	54.4	ug/L	0.445	8.81	20	
trans-1,3-Dichloropropene		50.0	49.5	ug/L	0.515	0.990	20	

CCV - Modified 03/05/2008
PDF File ID: 4831248
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927 Run Date: 06/23/2016 Sample ID: WG573990-02
Instrument ID: HPMS8 Run Time: 21:00 Method: 8260B
File ID: 8M413192 Analyst: FJB QC Key: DOD4
Workgroup (AAB#): WG573991 Cal ID: HPMS8 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
1,1-Dichloropropene		50.0	54.7	ug/L	0.431	9.40	20	
2-Hexanone		50.0	42.6	ug/L	0.0438	14.9	20	
Hexachlorobutadiene		50.0	49.1	ug/L	0.471	1.81	20	
Isopropylbenzene		50.0	49.8	ug/L	1.74	0.478	20	
p-Isopropyltoluene		50.0	48.6	ug/L	2.76	2.77	20	
4-Methyl-2-Pentanone		50.0	48.0	ug/L	0.0386	3.95	20	
Methylene Chloride		50.0	50.2	ug/L	0.258	0.364	20	
Naphthalene		50.0	43.5	ug/L	1.33	13.1	20	
n-Propylbenzene		50.0	48.3	ug/L	3.77	3.46	20	
Styrene		50.0	48.6	ug/L	1.05	2.85	20	
1,1,1,2-Tetrachloroethane		50.0	49.6	ug/L	0.402	0.773	20	
Tetrachloroethene		50.0	49.7	ug/L	0.331	0.579	20	
1,2,3-Trichlorobenzene		50.0	43.5	ug/L	0.754	13.0	20	
1,2,4-Trichlorobenzene		50.0	46.2	ug/L	0.949	7.59	20	
1,1,1-Trichloroethane		50.0	58.9	ug/L	0.609	17.8	20	
1,1,2-Trichloroethane		50.0	45.7	ug/L	0.224	8.57	20	
Trichloroethene		50.0	54.4	ug/L	0.298	8.88	20	
Trichlorofluoromethane		50.0	58.4	ug/L	0.630	16.7	20	
1,2,3-Trichloropropane		50.0	44.8	ug/L	0.139	10.3	20	
1,2,4-Trimethylbenzene		50.0	48.5	ug/L	2.84	2.95	20	
1,3,5-Trimethylbenzene		50.0	49.0	ug/L	2.78	2.08	20	
o-Xylene		50.0	47.1	ug/L	0.619	5.74	20	
m-,p-Xylene		100	96.4	ug/L	0.647	3.63	20	

* Exceeds %D Criteria

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

CCV - Modified 03/05/2008
PDF File ID: 4831248
Report generated 06/30/2016 15:02

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060927
Instrument ID: HPMS8
Workgroup (AAB#): WG573645

ICAL CCV Number: WG568561-08
CAL ID: HPMS8 - 13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568561-08	NA	NA	288225	539116	778260
Upper Limit	NA	NA	576450	1078232	1556520
Lower Limit	NA	NA	144113	269558	389130
L16060927-04	1.00	01	289525	517764	676177
L16060927-05	1.00	01	324131	585008	753725
WG573645-01	1.00	01	336359	605610	793415
WG573645-02	1.00	01	338876	608259	813093
WG573645-03	1.00	01	337211	606723	801570

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_ICAL - Modified 03/06/2008
PDF File ID: 4833868
Report generated 06/30/2016 15:03

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060927
Instrument ID: HPMS8
Workgroup (AAB#): WG573771

ICAL CCV Number: WG568561-08
CAL ID: HPMS8 - 13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568561-08	NA	NA	288225	539116	778260
Upper Limit	NA	NA	576450	1078232	1556520
Lower Limit	NA	NA	144113	269558	389130
L16060927-02	1.00	01	305295	540795	719947
WG573771-01	1.00	01	305326	554912	723354
WG573771-02	1.00	01	303556	551241	717279

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_ICAL - Modified 03/06/2008
PDF File ID: 4833868
Report generated 06/30/2016 15:03

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060927
Instrument ID: HPMS8
Workgroup (AAB#): WG573991

ICAL CCV Number: WG568561-08
CAL ID: HPMS8-13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568561-08	NA	NA	288225	539116	778260
Upper Limit	NA	NA	576450	1078232	1556520
Lower Limit	NA	NA	144113	269558	389130
L16060927-01	1.00	01	289917	516753	665575
L16060927-02	200	DL01	298222	542068	693727
L16060927-03	1.00	01	285769	514724	664173
WG573991-01	1.00	01	300577	547838	695599
WG573991-02	1.00	01	307359	541017	709070

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_ICAL - Modified 03/06/2008
PDF File ID: 4833868
Report generated 06/30/2016 15:03

Microbac Laboratories Inc.
INTERNAL STANDARD RETENTION TIME SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060927
Instrument ID: HPMS8
Workgroup (AAB#): WG573645

ICAL CCV Number: WG568561-08
CAL ID: HPMS8-13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568561-08	NA	NA	17.6	14.58	10.71
Upper Limit	NA	NA	18.1	15.08	11.21
Lower Limit	NA	NA	17.1	14.08	10.21
L16060927-04	1.00	01	17.6	14.57	10.71
L16060927-05	1.00	01	17.59	14.58	10.71
WG573645-01	1.00	01	17.6	14.57	10.71
WG573645-02	1.00	01	17.6	14.57	10.71
WG573645-03	1.00	01	17.6	14.57	10.71

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_RT_ICAL - Modified 03/06/2008
PDF File ID: 4833870
Report generated: 06/30/2016 15:03

Microbac Laboratories Inc.

INTERNAL STANDARD RETENTION TIME SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060927
Instrument ID: HPMS8
Workgroup (AAB#): WG573771

ICAL CCV Number: WG568561-08
CAL ID: HPMS8 - 13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568561-08	NA	NA	17.6	14.58	10.71
Upper Limit	NA	NA	18.1	15.08	11.21
Lower Limit	NA	NA	17.1	14.08	10.21
L16060927-02	1.00	01	17.6	14.57	10.71
WG573771-01	1.00	01	17.6	14.57	10.71
WG573771-02	1.00	01	17.6	14.57	10.71

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_RT_ICAL - Modified 03/06/2008
PDF File ID: 4833870
Report generated: 06/30/2016 15:03

Microbac Laboratories Inc.
INTERNAL STANDARD RETENTION TIME SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060927
Instrument ID: HPMS8
Workgroup (AAB#): WG573991

ICAL CCV Number: WG568561-08
CAL ID: HPMS8 - 13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568561-08	NA	NA	17.6	14.58	10.71
Upper Limit	NA	NA	18.1	15.08	11.21
Lower Limit	NA	NA	17.1	14.08	10.21
L16060927-01	1.00	01	17.6	14.58	10.71
L16060927-02	200	DL01	17.6	14.57	10.71
L16060927-03	1.00	01	17.6	14.57	10.71
WG573991-01	1.00	01	17.6	14.57	10.71
WG573991-02	1.00	01	17.6	14.57	10.71

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_RT_ICAL - Modified 03/06/2008
PDF File ID: 4833870
Report generated: 06/30/2016 15:03

2.1.1.3 Sample Data

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413199.D Vial: 10
Acq On : 24 Jun 2016 00:24 Operator: FJB
Sample : L16060927-01 C 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 24 00:46:08 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	665575	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	516753	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	289917	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.67	111	199022	27.8312 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 111.32%		
43) 1,2-Dichloroethane-d4	10.30	65	221801	28.9267 ug/L	-0.01
Spiked Amount 25.000	Range 80 - 120		Recovery = 115.72%		
58) Toluene-d8	12.68	98	701354	25.9334 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 103.72%		
80) p-Bromofluorobenzene	16.07	95	301623	25.1256 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 100.52%		

Target Compounds					Qvalue
3) Chloromethane	3.62	50	1785	0.2740 ug/L	80
13) Acetone	6.25	43	7239	9.6670 ug/L	93
20) Carbon Disulfide	7.29	76	4782	0.2123 ug/L #	74
28) Ethyl-Tert-Butyl ether	8.61	59	5727	0.3468 ug/L #	45
29) 2-Butanone	8.89	43	210	0.1909 ug/L #	47

(#) = qualifier out of range (m) = manual integration
8M413199.D 8260WT.M Mon Jun 27 12:41:49 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413199.D Vial: 10
Acq On : 24 Jun 2016 00:24 Operator: FJB
Sample : L16060927-01 C 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 24 0:46 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413199.D 8260WT.M Mon Jun 27 12:41:50 2016 Page 2



#3
Chloromethane
Concen: 0.27 ug/L
RT: 3.62 min Scan# 129
Delta R.T. 0.00 min
Lab File: 8M413199.D
Acq: 24 Jun 2016 00:24

Tgt Ion: 50 Resp: 1785

Ion	Ratio	Lower	Upper
50	100		
52	22.0	19.9	46.3





#13
Acetone
Concen: 9.67 ug/L
RT: 6.25 min Scan# 383
Delta R.T. 0.00 min
Lab File: 8M413199.D
Acq: 24 Jun 2016 00:24

Tgt Ion: 43 Resp: 7239

Ion	Ratio	Lower	Upper
43	100		
58	21.7	15.2	35.4





#20
Carbon Disulfide
Concen: 0.21 ug/L
RT: 7.29 min Scan# 484
Delta R.T. 0.00 min
Lab File: 8M413199.D
Acq: 24 Jun 2016 00:24

Tgt Ion: 76 Resp: 4782

Ion	Ratio	Lower	Upper
76	100		
78	0.0	5.6	13.0#





#28
Ethyl-Tert-Butyl ether
Concen: 0.35 ug/L
RT: 8.61 min Scan# 612
Delta R.T. -0.09 min
Lab File: 8M413199.D
Acq: 24 Jun 2016 00:24

Tgt Ion: 59 Resp: 5727

Ion	Ratio	Lower	Upper
59	100		
87	0.0	26.8	62.4#
57	10.1	17.4	40.6#





#29
2-Butanone
Concen: 0.19 ug/L
RT: 8.89 min Scan# 639
Delta R.T. 0.01 min
Lab File: 8M413199.D
Acq: 24 Jun 2016 00:24

Tgt Ion: 43 Resp: 210

Ion	Ratio	Lower	Upper
43	100		
72	0.0	16.4	38.4#



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413169.D Vial: 9
Acq On : 22 Jun 2016 18:40 Operator: FJB
Sample : L16060927-02 B 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 19:02:37 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	719947	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	540795	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	305295	25.00 ug/L	0.00

System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	215671	27.8817 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 111.52%		
43) 1,2-Dichloroethane-d4	10.30	65	237782	28.6689 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 114.68%		
58) Toluene-d8	12.68	98	745228	26.3306 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 105.32%		
80) p-Bromofluorobenzene	16.08	95	314218	24.8563 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 99.44%		

Target Compounds					Qvalue
4) Vinyl Chloride	3.84	62	49252	5.8917 ug/L	99
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	445801	60.9719 ug/L	96
13) Acetone	6.16	43	5498	6.7876 ug/L #	50
14) 1,1-Dichloroethene	6.47	61	117459	10.0895 ug/L	99
16) Dimethyl Sulfide	6.74	62	3777	0.9074 ug/L	92
19) Methylene Chloride	7.26	84	1895	0.2558 ug/L	88
20) Carbon Disulfide	7.30	76	3368	0.1382 ug/L #	74
23) trans-1,2-Dichloroethene	7.70	61	43691	4.0851 ug/L	95
27) 1,1-Dichloroethane	8.32	63	14036	1.0567 ug/L	94
32) cis-1,2-Dichloroethene	9.17	96	1920731	232.4198 ug/L	85
33) Chloroform	9.37	83	85451	5.4486 ug/L	99
36) Tetrahydrofuran	9.63	42	202	0.2577 ug/L #	24
39) Cyclohexane	9.94	56	1842	0.1786 ug/L #	76
45) 1,2-Dichloroethane	10.42	62	2158475	213.8341 ug/L	93
46) Benzene	10.47	78	47122	1.5934 ug/L #	1
47) Trichloroethene	11.21	130	16920395	2145.1934 ug/L	97
48) Methylcyclohexane	11.21	83	503714	42.0369 ug/L #	1
63) 1,1,2-Trichloroethane	13.18	97	4548	0.8585 ug/L	85
66) Tetrachloroethene	13.60	164	8783	1.2210 ug/L	97

(#) = qualifier out of range (m) = manual integration
8M413169.D 8260WT.M Mon Jun 27 12:40:10 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413169.D Vial: 9
Acq On : 22 Jun 2016 18:40 Operator: FJB
Sample : L16060927-02 B 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 19:02 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413169.D 8260WT.M Mon Jun 27 12:40:11 2016 Page 2



#4
Vinyl Chloride
Concen: 5.89 ug/L
RT: 3.84 min Scan# 151
Delta R.T. -0.00 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 62 Resp: 49252

Ion	Ratio	Lower	Upper
62	100		
64	31.9	19.3	45.1





#12
1,1,2-Trichloro-1,2,2-Trifluoroethane
Concen: 60.97 ug/L
RT: 6.16 min Scan# 375
Delta R.T. 0.01 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion:101 Resp: 445801

Ion	Ratio	Lower	Upper
101	100		
151	84.2	40.2	120.2





#13
Acetone
Concen: 6.79 ug/L
RT: 6.16 min Scan# 375
Delta R.T. -0.09 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 43 Resp: 5498

Ion	Ratio	Lower	Upper
43	100		
58	0.0	15.2	35.4#





#14
1,1-Dichloroethene
Concen: 10.09 ug/L
RT: 6.47 min Scan# 405
Delta R.T. 0.01 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 61 Resp: 117459

Ion	Ratio	Lower	Upper
61	100		
96	58.4	34.6	80.8





#16
Dimethyl Sulfide
Concen: 0.91 ug/L
RT: 6.74 min Scan# 431
Delta R.T. 0.01 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 62 Resp: 3777

Ion	Ratio	Lower	Upper
62	100		
47	81.0	53.2	124.0





#19
Methylene Chloride
Concen: 0.26 ug/L
RT: 7.26 min Scan# 481
Delta R.T. 0.01 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 84 Resp: 1895

Ion	Ratio	Lower	Upper
84	100		
49	88.8	60.5	141.3





#20
Carbon Disulfide
Concen: 0.14 ug/L
RT: 7.30 min Scan# 485
Delta R.T. 0.01 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 76 Resp: 3368

Ion	Ratio	Lower	Upper
76	100		
78	0.0	5.6	13.0#





#23
trans-1,2-Dichloroethene
Concen: 4.09 ug/L
RT: 7.70 min Scan# 524
Delta R.T. -0.00 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 61 Resp: 43691

Ion	Ratio	Lower	Upper
61	100		
96	65.9	41.8	97.4





#27
1,1-Dichloroethane
Concen: 1.06 ug/L
RT: 8.32 min Scan# 584
Delta R.T. -0.00 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 63 Resp: 14036

Ion	Ratio	Lower	Upper
63	100		
65	29.7	20.2	47.0
83	13.2	9.1	21.1





#32
cis-1,2-Dichloroethene
Concen: 232.42 ug/L
RT: 9.17 min Scan# 666
Delta R.T. -0.00 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 96 Resp: 1920731

Ion	Ratio	Lower	Upper
96	100		
61	132.6	91.5	213.5





#33
Chloroform
Concen: 5.45 ug/L
RT: 9.37 min Scan# 686
Delta R.T. -0.00 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 83 Resp: 85451

Ion	Ratio	Lower	Upper
83	100		
85	63.6	37.5	87.5





#36
Tetrahydrofuran
Concen: 0.26 ug/L
RT: 9.63 min Scan# 711
Delta R.T. -0.00 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 42 Resp: 202

Ion	Ratio	Lower	Upper
42	100		
72	0.0	33.8	78.8#
71	0.0	32.0	74.8#





#39
Cyclohexane
Concen: 0.18 ug/L
RT: 9.94 min Scan# 741
Delta R.T. 0.01 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 56 Resp: 1842

Ion	Ratio	Lower	Upper
56	100		
84	88.3	64.5	150.5
69	13.0	22.0	51.4#





#45
1,2-Dichloroethane
Concen: 213.83 ug/L
RT: 10.42 min Scan# 787
Delta R.T. -0.00 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 62 Resp: 2158475

Ion	Ratio	Lower	Upper
62	100		
64	33.2	19.6	45.6
49	18.7	15.5	36.1





#46
Benzene
Concen: 1.59 ug/L
RT: 10.47 min Scan# 792
Delta R.T. 0.01 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 78 Resp: 47122

Ion	Ratio	Lower	Upper
78	100		
52	16.3	8.4	19.6
51	286.4	10.9	25.5#





#47
Trichloroethene
Concen: 2145.19 ug/L
RT: 11.21 min Scan# 864
Delta R.T. -0.00 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion:130 Resp:16920395

Ion	Ratio	Lower	Upper
130	100		
132	98.2	58.0	135.2
95	103.0	64.9	151.5





#48
Methylcyclohexane
Concen: 42.04 ug/L
RT: 11.21 min Scan# 864
Delta R.T. -0.09 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 83 Resp: 503714

Ion	Ratio	Lower	Upper
83	100		
55	0.1	40.9	95.3#
98	199.2	28.1	65.7#





#63
1,1,2-Trichloroethane
Concen: 0.86 ug/L
RT: 13.18 min Scan# 1054
Delta R.T. -0.00 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion: 97 Resp: 4548

Ion	Ratio	Lower	Upper
97	100		
83	70.9	50.9	118.9





#66
Tetrachloroethene
Concen: 1.22 ug/L
RT: 13.60 min Scan# 1095
Delta R.T. -0.00 min
Lab File: 8M413169.D
Acq: 22 Jun 2016 18:40

Tgt Ion:164 Resp: 8783

Ion	Ratio	Lower	Upper
164	100		
129	98.0	56.6	132.2
94	70.5	40.7	95.1



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413196.D Vial: 7
Acq On : 23 Jun 2016 22:56 Operator: FJB
Sample : L16060927-02 200X C 826-LOW D1 Inst : HPMS8
Misc : 1,200 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 23 23:18:55 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	693727	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	542068	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	298222	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.67	111	209484	28.1054 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 112.44%		
43) 1,2-Dichloroethane-d4	10.30	65	232136	29.0460 ug/L	-0.01
Spiked Amount 25.000	Range 80 - 120		Recovery = 116.20%		
58) Toluene-d8	12.68	98	719106	25.3480 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 101.40%		
80) p-Bromofluorobenzene	16.07	95	310834	25.1718 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 100.68%		

Target Compounds

					Qvalue
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	1124	0.1595 ug/L	69
19) Methylene Chloride	7.25	84	1850	0.2591 ug/L	94
20) Carbon Disulfide	7.29	76	13434	0.5723 ug/L #	80
32) cis-1,2-Dichloroethene	9.17	96	9050	1.1365 ug/L	83
45) 1,2-Dichloroethane	10.42	62	10312	1.0602 ug/L	91
47) Trichloroethene	11.22	130	145308	19.1187 ug/L	99
48) Methylcyclohexane	11.21	83	1919	0.1662 ug/L #	1
51) 1,4-Dioxane	11.70	88	888	27.8668 ug/L #	26
98) 1,2,4-Trichlorobenzene	20.28	180	1884	0.1538 ug/L #	76
100) Naphthalene	20.64	128	6385	0.3508 ug/L	99
101) 1,2,3-Trichlorobenzene	20.96	180	1289	0.1247 ug/L #	76

(#) = qualifier out of range (m) = manual integration
8M413196.D 8260WT.M Mon Jun 27 12:41:35 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413196.D Vial: 7
Acq On : 23 Jun 2016 22:56 Operator: FJB
Sample : L16060927-02 200X C 826-LOW D1 Inst : HPMS8
Misc : 1,200 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 23 23:18 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413196.D 8260WT.M Mon Jun 27 12:41:36 2016 Page 2



#12
1,1,2-Trichloro-1,2,2-Trifluoroethane
Concen: 0.16 ug/L
RT: 6.15 min Scan# 374
Delta R.T. -0.00 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion:101 Resp: 1124

Ion	Ratio	Lower	Upper
101	100		
151	52.9	40.2	120.2





#19
Methylene Chloride
Concen: 0.26 ug/L
RT: 7.25 min Scan# 480
Delta R.T. -0.00 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion: 84 Resp: 1850

Ion	Ratio	Lower	Upper
84	100		
49	95.2	60.5	141.3





#20
Carbon Disulfide
Concen: 0.57 ug/L
RT: 7.29 min Scan# 484
Delta R.T. -0.00 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion: 76 Resp: 13434

Ion	Ratio	Lower	Upper
76	100		
78	1.9	5.6	13.0#





#32
cis-1,2-Dichloroethene
Concen: 1.14 ug/L
RT: 9.17 min Scan# 666
Delta R.T. -0.00 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion: 96 Resp: 9050

Ion	Ratio	Lower	Upper
96	100		
61	130.5	91.5	213.5





#45
1,2-Dichloroethane
Concen: 1.06 ug/L
RT: 10.42 min Scan# 787
Delta R.T. -0.00 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion: 62 Resp: 10312

Ion	Ratio	Lower	Upper
62	100		
64	31.6	19.6	45.6
49	16.4	15.5	36.1





#47
Trichloroethene
Concen: 19.12 ug/L
RT: 11.22 min Scan# 864
Delta R.T. -0.00 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion:130 Resp: 145308

Ion	Ratio	Lower	Upper
130	100		
132	97.6	58.0	135.2
95	108.0	64.9	151.5





#48
Methylcyclohexane
Concen: 0.17 ug/L
RT: 11.21 min Scan# 863
Delta R.T. -0.09 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion: 83 Resp: 1919

Ion	Ratio	Lower	Upper
83	100		
55	0.0	40.9	95.3#
98	246.0	28.1	65.7#





#51
1,4-Dioxane
Concen: 27.87 ug/L
RT: 11.70 min Scan# 911
Delta R.T. -0.01 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion: 88 Resp: 888

Ion	Ratio	Lower	Upper
88	100		
58	0.0	31.4	73.2#





#98
1,2,4-Trichlorobenzene
Concen: 0.15 ug/L
RT: 20.28 min Scan# 1741
Delta R.T. -0.00 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion:180 Resp: 1884

Ion	Ratio	Lower	Upper
180	100		
182	74.8	56.8	132.4
145	11.2	18.4	42.8#





#100
Naphthalene
Concen: 0.35 ug/L
RT: 20.64 min Scan# 1776
Delta R.T. -0.01 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion:128 Resp: 6385

Ion	Ratio	Lower	Upper
128	100		
127	12.8	7.8	18.2





#101
1,2,3-Trichlorobenzene
Concen: 0.12 ug/L
RT: 20.96 min Scan# 1807
Delta R.T. -0.00 min
Lab File: 8M413196.D
Acq: 23 Jun 2016 22:56

Tgt Ion:180 Resp: 1289

Ion	Ratio	Lower	Upper
180	100		
182	84.8	57.8	135.0
145	0.0	20.1	46.9#



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413200.D Vial: 11
Acq On : 24 Jun 2016 00:53 Operator: FJB
Sample : L16060927-03 C 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 24 01:15:15 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	664173	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	514724	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	285769	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.67	111	201374	28.2196 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 112.88%		
43) 1,2-Dichloroethane-d4	10.31	65	228249	29.8305 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 119.32%		
58) Toluene-d8	12.68	98	698077	25.9140 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 103.64%		
80) p-Bromofluorobenzene	16.07	95	305251	25.7969 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 103.20%		

Target Compounds

					Qvalue
2) Dichlorodifluoromethane	3.08	85	165260	14.8408 ug/L	99
3) Chloromethane	3.61	50	1709	0.2629 ug/L	90
8) Trichlorofluoromethane	5.39	101	6942	0.4838 ug/L #	1
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	1575828	233.6238 ug/L	94
13) Acetone	6.16	43	19869	26.5893 ug/L #	70
20) Carbon Disulfide	7.29	76	13455	0.5987 ug/L	97
28) Ethyl-Tert-Butyl ether	8.62	59	5420	0.3289 ug/L #	47
33) Chloroform	9.39	83	3618	0.2501 ug/L	89
35) Bromochloromethane	9.60	130	1515	0.3875 ug/L	94
52) Dibromomethane	11.81	93	1819	0.5060 ug/L	84

(#) = qualifier out of range (m) = manual integration
8M413200.D 8260WT.M Mon Jun 27 12:41:53 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413200.D Vial: 11
Acq On : 24 Jun 2016 00:53 Operator: FJB
Sample : L16060927-03 C 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 24 1:15 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration





#2
Dichlorodifluoromethane
Concen: 14.84 ug/L
RT: 3.08 min Scan# 77
Delta R.T. -0.08 min
Lab File: 8M413200.D
Acq: 24 Jun 2016 00:53

Tgt Ion: 85 Resp: 165260

Ion	Ratio	Lower	Upper
85	100		
87	32.0	19.6	45.8





#3
Chloromethane
Concen: 0.26 ug/L
RT: 3.61 min Scan# 128
Delta R.T. -0.01 min
Lab File: 8M413200.D
Acq: 24 Jun 2016 00:53

Tgt Ion: 50 Resp: 1709

Ion	Ratio	Lower	Upper
50	100		
52	27.6	19.9	46.3





#8
Trichlorofluoromethane
Concen: 0.48 ug/L
RT: 5.39 min Scan# 300
Delta R.T. 0.01 min
Lab File: 8M413200.D
Acq: 24 Jun 2016 00:53

Tgt Ion:101 Resp: 6942

Ion	Ratio	Lower	Upper
101	100		
103	52.3	39.2	91.6
66	442.0	7.0	16.4#





#12
1,1,2-Trichloro-1,2,2-Trifluoroethane
Concen: 233.62 ug/L
RT: 6.16 min Scan# 375
Delta R.T. 0.01 min
Lab File: 8M413200.D
Acq: 24 Jun 2016 00:53

Tgt Ion:101 Resp: 1575828

Ion	Ratio	Lower	Upper
101	100		
151	85.1	40.2	120.2





#13
Acetone
Concen: 26.59 ug/L
RT: 6.16 min Scan# 375
Delta R.T. -0.08 min
Lab File: 8M413200.D
Acq: 24 Jun 2016 00:53

Tgt Ion: 43 Resp: 19869

Ion	Ratio	Lower	Upper
43	100		
58	9.9	15.2	35.4#





#20
Carbon Disulfide
Concen: 0.60 ug/L
RT: 7.29 min Scan# 484
Delta R.T. -0.00 min
Lab File: 8M413200.D
Acq: 24 Jun 2016 00:53

Tgt Ion: 76 Resp: 13455

Ion	Ratio	Lower	Upper
76	100		
78	8.3	5.6	13.0





#28
Ethyl-Tert-Butyl ether
Concen: 0.33 ug/L
RT: 8.62 min Scan# 613
Delta R.T. -0.08 min
Lab File: 8M413200.D
Acq: 24 Jun 2016 00:53

Tgt Ion: 59 Resp: 5420

Ion	Ratio	Lower	Upper
59	100		
87	0.0	26.8	62.4#
57	13.9	17.4	40.6#





#33
Chloroform
Concen: 0.25 ug/L
RT: 9.39 min Scan# 687
Delta R.T. 0.01 min
Lab File: 8M413200.D
Acq: 24 Jun 2016 00:53

Tgt Ion: 83 Resp: 3618

Ion	Ratio	Lower	Upper
83	100		
85	54.3	37.5	87.5





#35
Bromochloromethane
Concen: 0.39 ug/L
RT: 9.60 min Scan# 708
Delta R.T. -0.00 min
Lab File: 8M413200.D
Acq: 24 Jun 2016 00:53

Tgt Ion:130 Resp: 1515

Ion	Ratio	Lower	Upper
130	100		
128	72.5	46.1	107.5
49	97.8	63.3	147.7





#52
Dibromomethane
Concen: 0.51 ug/L
RT: 11.81 min Scan# 921
Delta R.T. -0.01 min
Lab File: 8M413200.D
Acq: 24 Jun 2016 00:53

Tgt Ion: 93 Resp: 1819

Ion	Ratio	Lower	Upper
93	100		
95	78.0	49.7	116.1
174	78.7	62.3	145.3



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413145.D Vial: 22
Acq On : 22 Jun 2016 3:34 Operator: FJB
Sample : L16060927-04 A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 03:56:14 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	676177	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	517764	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	289525	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	202889	27.9271 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 111.72%		
43) 1,2-Dichloroethane-d4	10.30	65	222170	28.5205 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 114.08%		
58) Toluene-d8	12.68	98	707274	26.1012 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 104.40%		
80) p-Bromofluorobenzene	16.08	95	299500	24.9825 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 99.92%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration
8M413145.D 8260WT.M Fri Jun 24 17:21:15 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413145.D Vial: 22
Acq On : 22 Jun 2016 3:34 Operator: FJB
Sample : L16060927-04 A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 3:56 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413145.D 8260WT.M Fri Jun 24 17:21:16 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413135.D Vial: 12
Acq On : 21 Jun 2016 22:44 Operator: FJB
Sample : L16060927-05 A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 23:06:49 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	753725	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	585008	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.59	152	324131	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.67	111	224626	27.7380 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 110.96%		
43) 1,2-Dichloroethane-d4	10.31	65	243766	28.0732 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 112.28%		
58) Toluene-d8	12.69	98	790206	25.8097 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 103.24%		
80) p-Bromofluorobenzene	16.07	95	337050	25.1130 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 100.44%		

Target Compounds					Qvalue
36) Tetrahydrofuran	9.63	42	1126	1.3722 ug/L #	52

(#) = qualifier out of range (m) = manual integration
8M413135.D 8260WT.M Fri Jun 24 17:20:31 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413135.D Vial: 12
Acq On : 21 Jun 2016 22:44 Operator: FJB
Sample : L16060927-05 A 826-LOW Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 23:06 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration





#36
Tetrahydrofuran
Concen: 1.37 ug/L
RT: 9.63 min Scan# 710
Delta R.T. -0.01 min
Lab File: 8M413135.D
Acq: 21 Jun 2016 22:44

Tgt Ion: 42 Resp: 1126

Ion	Ratio	Lower	Upper
42	100		
72	73.0	33.8	78.8
71	0.0	32.0	74.8#



2.1.1.4 Standards Data

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\2\DATA\051316\8M412220.D Vial: 3
Acq On : 13 May 2016 13:29 Operator: TMB
Sample : WG568561-02 0.3ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 17 11:40:52 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	823748	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	571926	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	299321	25.00 ug/L	0.00

System Monitoring Compounds					
37) Dibromofluoromethane	0.00	111	0	0.0000 ug/L	
Spiked Amount 25.000	Range 86 - 118		Recovery	= 0.00%#	
43) 1,2-Dichloroethane-d4	0.00	65	0	0.0000 ug/L	
Spiked Amount 25.000	Range 80 - 120		Recovery	= 0.00%#	
58) Toluene-d8	12.68	98	567	0.0189 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery	= 0.08%#	
80) p-Bromofluorobenzene	16.08	95	313	0.0253 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery	= 0.12%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.17	85	3836	0.2778 ug/L	91
3) Chloromethane	3.62	50	3658	0.4537 ug/L	95
4) Vinyl Chloride	3.85	62	2516	0.2630 ug/L #	58
5) 1,3-Butadiene	3.91	54	2079	0.2592 ug/L	71
6) Bromomethane	4.73	94	1152	0.6530 ug/L #	55
7) Chloroethane	4.90	64	1112	0.2113 ug/L #	72
8) Trichlorofluoromethane	5.35	101	2706	0.1520 ug/L #	53
10) Isoprene	5.93	67	2147	0.1987 ug/L	85
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	1746	0.2087 ug/L	97
14) 1,1-Dichloroethene	6.48	61	3332	0.2501 ug/L	75
16) Dimethyl Sulfide	6.73	62	1099	0.2308 ug/L	95
19) Methylene Chloride	7.26	84	3414	0.4028 ug/L	88
20) Carbon Disulfide	7.30	76	10739	0.3853 ug/L	97
22) Methyl Tert Butyl Ether	7.47	73	4927	0.2686 ug/L #	60
23) trans-1,2-Dichloroethene	7.70	61	3879	0.3170 ug/L	98
24) n-Hexane	7.78	57	2794	0.3075 ug/L	85
26) Vinyl Acetate	8.30	43	1973	0.2259 ug/L #	70
27) 1,1-Dichloroethane	8.33	63	4881	0.3212 ug/L #	88
31) 2,2-Dichloropropane	9.11	77	4966	0.3101 ug/L	89
32) cis-1,2-Dichloroethene	9.17	96	2958	0.3128 ug/L	87
33) Chloroform	9.37	83	5564	0.3101 ug/L	95
34) 1-Bromopropane	9.52	122	210	0.1551 ug/L #	1
35) Bromochloromethane	9.60	130	1386	0.2858 ug/L	76
36) Tetrahydrofuran	9.62	42	1849	2.0617 ug/L #	60
38) 1,1,1-Trichloroethane	9.91	97	4978	0.2921 ug/L	95
39) Cyclohexane	9.93	56	3120	0.2644 ug/L #	85
40) 1,1-Dichloropropene	10.11	75	3779	0.2909 ug/L	97
42) Carbon Tetrachloride	10.24	117	4835	0.3194 ug/L #	74
45) 1,2-Dichloroethane	10.43	62	3295	0.2853 ug/L #	81
46) Benzene	10.46	78	11279	0.3333 ug/L	93
47) Trichloroethene	11.20	130	2410	0.2670 ug/L	97
48) Methylcyclohexane	11.31	83	3936	0.2871 ug/L	84
49) 1,2-Dichloropropane	11.42	63	2136	0.2836 ug/L	88
50) Bromodichloromethane	11.73	83	3767	0.2845 ug/L #	91
52) Dibromomethane	11.82	93	700	0.1570 ug/L #	59
53) 2-Chloroethyl Vinyl Ether	12.02	63	847	0.2616 ug/L #	47
55) cis-1,3-Dichloropropene	12.36	75	4763	0.3535 ug/L	84
56) Dimethyl Disulfide	12.63	79	1647	0.2290 ug/L	82
59) Toluene	12.78	91	11513	0.3243 ug/L	92
60) Ethyl Methacrylate	12.89	69	1631	0.2392 ug/L #	64

(#) = qualifier out of range (m) = manual integration
8M412220.D 8260WT.M Tue May 17 11:40:54 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\2\DATA\051316\8M412220.D Vial: 3
Acq On : 13 May 2016 13:29 Operator: TMB
Sample : WG568561-02 0.3ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 17 11:40:52 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
62)	trans-1,3-Dichloropropene	12.96	75	3626	0.3047 ug/L		82
63)	1,1,2-Trichloroethane	13.18	97	1409	0.2515 ug/L		98
65)	1,3-Dichloropropane	13.49	76	2584	0.2590 ug/L		93
66)	Tetrachloroethene	13.61	164	2678	0.3520 ug/L	#	75
67)	Dibromochloromethane	13.87	129	2126	0.2599 ug/L		79
68)	1,2-Dibromoethane	14.12	107	1895	0.3394 ug/L		91
69)	1-Chlorohexane	14.21	91	3835	0.3164 ug/L		99
70)	Chlorobenzene	14.62	112	7968	0.3328 ug/L		88
71)	1,1,1,2-Tetrachloroethane	14.67	131	2872	0.3096 ug/L		87
72)	Ethylbenzene	14.64	106	3844	0.2934 ug/L		81
73)	m-,p-Xylene	14.75	106	10447	0.6799 ug/L		96
74)	o-Xylene	15.31	106	4577	0.3047 ug/L		97
75)	Styrene	15.35	104	7545	0.3062 ug/L		95
76)	Bromoform	15.84	173	1301	0.2681 ug/L	#	51
77)	Isopropylbenzene	15.74	105	13119	0.3287 ug/L		90
79)	1,1,2,2-Tetrachloroethane	15.96	83	1714	0.2853 ug/L	#	85
83)	n-Propylbenzene	16.24	91	15676	0.3357 ug/L		92
84)	Bromobenzene	16.38	156	2859	0.2977 ug/L		90
85)	1,3,5-Trimethylbenzene	16.43	105	10873	0.3196 ug/L		99
86)	2-Chlorotoluene	16.52	91	11133	0.3415 ug/L		96
87)	4-Chlorotoluene	16.56	91	10305	0.3499 ug/L		97
88)	a-Methylstyrene	16.83	118	4090	0.2534 ug/L		86
89)	tert-Butylbenzene	16.90	134	1734	0.2606 ug/L		72
90)	1,2,4-Trimethylbenzene	16.95	105	10734	0.3069 ug/L		89
91)	sec-Butylbenzene	17.16	105	13749	0.3436 ug/L		97
92)	p-Isopropyltoluene	17.32	119	10847	0.3192 ug/L		95
93)	1,3-Dichlorobenzene	17.52	146	6023	0.3257 ug/L		99
94)	1,4-Dichlorobenzene	17.65	146	6152	0.3299 ug/L	#	60
95)	n-Butylbenzene	17.85	91	11161	0.3323 ug/L		97
96)	1,2-Dichlorobenzene	18.14	146	5711	0.3472 ug/L		92
98)	1,2,4-Trichlorobenzene	20.29	180	4203	0.3418 ug/L		91
99)	Hexachlorobutadiene	20.43	225	1894	0.3297 ug/L		84
100)	Naphthalene	20.65	128	6039	0.3306 ug/L		91
101)	1,2,3-Trichlorobenzene	20.96	180	3279	0.3161 ug/L		86

(#) = qualifier out of range (m) = manual integration

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412220.D Vial: 3
Acq On : 13 May 2016 13:29 Operator: TMB
Sample : WG568561-02 0.3ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:03:01 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	-1.0000	0.2778	0.0	100	0.00
3	P	Chloromethane	-1.0000	0.4537	0.0	100	0.00
4	C	Vinyl Chloride	-1.0000	0.2630	0.0	100	0.00
5	T	1,3-Butadiene	-1.0000	0.2592	0.0	100	0.03
6	T	Bromomethane	-1.0000	0.9911	0.0	100	0.00
7	T	Chloroethane	-1.0000	0.2113	0.0	100	0.00
8	T	Trichlorofluoromethane	-1.0000	0.1520	0.0	100	-0.02
9	T	Diethyl ether	-1.0000	0.0000	0.0	0	-5.90#
10	T	Isoprene	-1.0000	0.1987	0.0	100	0.00
11	T	Acrolein	-1.0000	0.0000	0.0	0	-6.13#
12	T	1,1,2-Trichloro-1,2,2-Trifl	-1.0000	0.2087	0.0	100	0.00
13	T	Acetone	-1.0000	0.0000	0.0	0	-6.25#
14	C	1,1-Dichloroethene	-1.0000	0.2501	0.0	100	0.02
15	T	Tert-Butyl Alcohol	-1.0000	0.0000	0.0	0	-6.58#
16	T	Dimethyl Sulfide	-1.0000	0.2308	0.0	100	0.00
17	T	Iodomethane	-1.0000	0.0000	0.0	0	-6.98#
18	T	Methyl acetate	-1.0000	0.0000	0.0	0	-7.00#
19	T	Methylene Chloride	-1.0000	0.4028	0.0	100	0.00
20	T	Carbon Disulfide	-1.0000	0.3853	0.0	100	0.00
21	T	Acrylonitrile	-1.0000	0.0000	0.0	0	-7.43#
22	T	Methyl Tert Butyl Ether	-1.0000	0.2686	0.0	100	0.00
23	T	trans-1,2-Dichloroethene	-1.0000	0.3170	0.0	100	0.00
24	T	n-Hexane	-1.0000	0.3075	0.0	100	0.00
25	T	Diisopropyl ether	-1.0000	0.0000	0.0	0	-8.13#
26	T	Vinyl Acetate	-1.0000	0.2259	0.0	100	0.00
27	P	1,1-Dichloroethane	-1.0000	0.3212	0.0	100	0.00
28	T	Ethyl-Tert-Butyl ether	-1.0000	0.0000	0.0	0	-8.71#
29	T	2-Butanone	-1.0000	0.0000	0.0	0	-8.88#
30	T	Propionitrile	-1.0000	0.0000	0.0	0	-8.98#
31	T	2,2-Dichloropropane	-1.0000	0.3101	0.0	100	0.00
32	T	cis-1,2-Dichloroethene	-1.0000	0.3128	0.0	100	0.00
33	C	Chloroform	0.3000	0.3101	-3.4	100	0.00
34		1-Bromopropane	-1.0000	0.1551	0.0	100	0.00
35	T	Bromochloromethane	-1.0000	0.2858	0.0	100	0.00
36	T	Tetrahydrofuran	-1.0000	2.0617	0.0	100	-0.01
37	S	Dibromofluoromethane	-1.0000	0.0000	0.0	0	-9.67#
38	T	1,1,1-Trichloroethane	-1.0000	0.2921	0.0	100	0.00
39	T	Cyclohexane	-1.0000	0.2644	0.0	100	0.00
40	T	1,1-Dichloropropene	-1.0000	0.2909	0.0	100	0.00
41	T	Tert-Amyl-Methyl ether	-1.0000	0.0000	0.0	0	-10.21#
42	T	Carbon Tetrachloride	-1.0000	0.3194	0.0	100	0.00
43	S	1,2-Dichloroethane-d4	-1.0000	0.0000	0.0	0	-10.31#
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	-1.0000	0.2853	0.0	100	0.00
46	T	Benzene	-1.0000	0.3333	0.0	100	0.00
47	T	Trichloroethene	-1.0000	0.2670	0.0	100	-0.01
48	T	Methylcyclohexane	-1.0000	0.2871	0.0	100	0.00
49	C	1,2-Dichloropropane	-1.0000	0.2836	0.0	100	0.00
50	T	Bromodichloromethane	-1.0000	0.2845	0.0	100	0.00
51	T	1,4-Dioxane	-1.0000	0.0000	0.0	0	-11.71#
52	T	Dibromomethane	-1.0000	0.1570	0.0	100	0.00
53	T	2-Chloroethyl Vinyl Ether	-1.0000	0.2616	0.0	100	0.00
54	T	4-Methyl-2-Pentanone	-1.0000	0.0000	0.0	0	-12.05#

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412220.D Vial: 3
Acq On : 13 May 2016 13:29 Operator: TMB
Sample : WG568561-02 0.3ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:03:01 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	-1.0000	0.3535	0.0	100	0.00
56	T	Dimethyl Disulfide	-1.0000	0.2290	0.0	100	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	-1.0000	0.0189	0.0	100	0.00
59	C	Toluene	-1.0000	0.3243	0.0	100	0.00
60	T	Ethyl Methacrylate	-1.0000	0.2392	0.0	100	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	-1.0000	0.3047	0.0	100	0.00
63	T	1,1,2-Trichloroethane	-1.0000	0.2515	0.0	100	0.00
64	T	2-Hexanone	-1.0000	0.0000	0.0	0	-13.12#
65	T	1,3-Dichloropropane	-1.0000	0.2590	0.0	100	0.00
66	T	Tetrachloroethene	-1.0000	0.3520	0.0	100	0.00
67	T	Dibromochloromethane	-1.0000	0.2599	0.0	100	0.00
68	T	1,2-Dibromoethane	-1.0000	0.3394	0.0	100	0.00
69	T	1-Chlorohexane	-1.0000	0.3164	0.0	100	0.00
70	P	Chlorobenzene	-1.0000	0.3328	0.0	100	0.00
71	T	1,1,1,2-Tetrachloroethane	-1.0000	0.3096	0.0	100	0.00
72	C	Ethylbenzene	-1.0000	0.2934	0.0	100	-0.01
73	T	m-,p-Xylene	-1.0000	0.6799	0.0	100	0.00
74	T	o-Xylene	-1.0000	0.3047	0.0	100	0.00
75	T	Styrene	-1.0000	0.3062	0.0	100	0.00
76	P	Bromoform	-1.0000	0.2681	0.0	100	0.00
77	T	Isopropylbenzene	-1.0000	0.3287	0.0	100	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
79	P	1,1,2,2-Tetrachloroethane	-1.0000	0.2853	0.0	100	0.00
80	S	p-Bromofluorobenzene	-1.0000	0.0253	0.0	100	0.00
81	T	1,2,3-Trichloropropane	-1.0000	0.1103	0.0	100	0.00
82	T	trans-1,4-Dichloro-2-Butene	-1.0000	0.0000	0.0	0	-16.19#
83	T	n-Propylbenzene	-1.0000	0.3357	0.0	100	0.00
84	T	Bromobenzene	0.3000	0.2977	0.8	100	0.00
85	T	1,3,5-Trimethylbenzene	-1.0000	0.3196	0.0	100	0.00
86	T	2-Chlorotoluene	-1.0000	0.3415	0.0	100	0.00
87	T	4-Chlorotoluene	-1.0000	0.3499	0.0	100	0.00
88	T	a-Methylstyrene	-1.0000	0.2534	0.0	100	0.00
89	T	tert-Butylbenzene	-1.0000	0.2606	0.0	100	0.00
90	T	1,2,4-Trimethylbenzene	-1.0000	0.3069	0.0	100	0.00
91	T	sec-Butylbenzene	-1.0000	0.3436	0.0	100	0.00
92	T	p-Isopropyltoluene	-1.0000	0.3192	0.0	100	0.00
93	T	1,3-Dichlorobenzene	-1.0000	0.3257	0.0	100	0.00
94	T	1,4-Dichlorobenzene	0.3000	0.3299	-10.0	100	0.00
95	T	n-Butylbenzene	-1.0000	0.3323	0.0	100	0.00
96	T	1,2-Dichlorobenzene	0.3000	0.3472	-15.7	100	0.00
97	T	1,2-Dibromo-3-Chloropropane	-1.0000	0.0000	0.0	0	-19.13#
98	T	1,2,4-Trichlorobenzene	-1.0000	0.3418	0.0	100	0.00
99	T	Hexachlorobutadiene	-1.0000	0.3297	0.0	100	0.00
100	T	Naphthalene	-1.0000	0.3306	0.0	100	0.00
101	T	1,2,3-Trichlorobenzene	0.3000	0.3161	-5.4	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412220.D 8260WT.M Sat May 14 18:03:34 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412221.D Vial: 4
Acq On : 13 May 2016 13:58 Operator: TMB
Sample : WG568561-03 0.4ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:20:36 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	801173	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	557498	25.00 ug/L	0.01
78) 1,4-Dichlorobenzene-d4	17.60	152	293108	25.00 ug/L	0.01

System Monitoring Compounds					
37) Dibromofluoromethane	0.00	111	0	0.0000 ug/L	
Spiked Amount 25.000	Range 86 - 118		Recovery	= 0.00%#	
43) 1,2-Dichloroethane-d4	0.00	65	0	0.0000 ug/L	
Spiked Amount 25.000	Range 80 - 120		Recovery	= 0.00%#	
58) Toluene-d8	0.00	98	0	0.0000 ug/L	
Spiked Amount 25.000	Range 88 - 110		Recovery	= 0.00%#	
80) p-Bromofluorobenzene	16.08	95	343	0.0286 ug/L	0.01
Spiked Amount 25.000	Range 86 - 115		Recovery	= 0.12%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	5478	0.3262 ug/L	79
3) Chloromethane	3.63	50	3930	0.2484 ug/L	85
4) Vinyl Chloride	3.84	62	3512	0.3288 ug/L	97
5) 1,3-Butadiene	3.90	54	2479	0.2375 ug/L #	75
6) Bromomethane	4.73	94	1343	0.9497 ug/L	85
7) Chloroethane	4.90	64	2219	0.3174 ug/L #	43
8) Trichlorofluoromethane	5.36	101	6974	0.3289 ug/L #	93
10) Isoprene	5.94	67	4288	0.3952 ug/L #	62
12) 1,1,2-Trichloro-1,2,2-Trif	6.17	101	3168	0.3263 ug/L #	40
14) 1,1-Dichloroethene	6.47	61	5431	0.2882 ug/L	92
16) Dimethyl Sulfide	6.74	62	1708	0.2453 ug/L	64
19) Methylene Chloride	7.24	84	3785	0.4239 ug/L #	43
20) Carbon Disulfide	7.29	76	12437	0.4565 ug/L	97
22) Methyl Tert Butyl Ether	7.47	73	6580	0.3336 ug/L #	76
23) trans-1,2-Dichloroethene	7.71	61	4990	0.2928 ug/L	96
24) n-Hexane	7.79	57	3110	0.2011 ug/L	99
26) Vinyl Acetate	8.31	43	2753	0.1566 ug/L	100
27) 1,1-Dichloroethane	8.32	63	5655	0.2774 ug/L #	92
31) 2,2-Dichloropropane	9.10	77	6035	0.3360 ug/L	82
32) cis-1,2-Dichloroethene	9.16	96	3837	0.3762 ug/L	74
33) Chloroform	9.37	83	7267	0.3842 ug/L	100
34) 1-Bromopropane	9.50	122	234	0.1456 ug/L #	1
35) Bromochloromethane	9.60	130	1674	0.3137 ug/L	62
36) Tetrahydrofuran	9.62	42	447	0.2146 ug/L #	41
38) 1,1,1-Trichloroethane	9.91	97	6304	0.3305 ug/L	90
39) Cyclohexane	9.94	56	4658	0.2238 ug/L #	67
40) 1,1-Dichloropropene	10.11	75	5172	0.3768 ug/L	80
42) Carbon Tetrachloride	10.25	117	5692	0.3172 ug/L	95
45) 1,2-Dichloroethane	10.43	62	4443	0.2796 ug/L #	63
46) Benzene	10.46	78	14456	0.4134 ug/L #	77
47) Trichloroethene	11.21	130	3659	0.3626 ug/L	98
48) Methylcyclohexane	11.29	83	5025	0.3562 ug/L	84
49) 1,2-Dichloropropane	11.42	63	3060	0.3007 ug/L	98
50) Bromodichloromethane	11.73	83	4734	0.3343 ug/L #	93
52) Dibromomethane	11.82	93	1805	0.3767 ug/L	84
55) cis-1,3-Dichloropropene	12.36	75	5008	0.3514 ug/L	90
56) Dimethyl Disulfide	12.63	79	1921	0.2573 ug/L	85
59) Toluene	12.78	91	14742	0.4416 ug/L	98
60) Ethyl Methacrylate	12.90	69	2062	0.3292 ug/L	66
62) trans-1,3-Dichloropropene	12.96	75	3693	0.3171 ug/L	100

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412221.D Vial: 4
Acq On : 13 May 2016 13:58 Operator: TMB
Sample : WG568561-03 0.4ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:20:36 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
63)	1,1,2-Trichloroethane	13.18	97	2326	0.4391 ug/L		70
65)	1,3-Dichloropropane	13.49	76	3163	0.3219 ug/L		88
66)	Tetrachloroethene	13.61	164	3164	0.4185 ug/L		93
67)	Dibromochloromethane	13.87	129	3113	0.3727 ug/L		86
68)	1,2-Dibromoethane	14.12	107	2076	0.3760 ug/L		90
69)	1-Chlorohexane	14.21	91	4746	0.4174 ug/L		80
70)	Chlorobenzene	14.62	112	9180	0.3960 ug/L		98
71)	1,1,1,2-Tetrachloroethane	14.66	131	3501	0.3876 ug/L		95
72)	Ethylbenzene	14.65	106	5348	0.4066 ug/L		84
73)	m-,p-Xylene	14.75	106	13269	0.8757 ug/L		89
74)	o-Xylene	15.31	106	6019	0.3943 ug/L		97
75)	Styrene	15.35	104	9370	0.3849 ug/L		95
76)	Bromoform	15.83	173	1198	0.2500 ug/L	#	76
77)	Isopropylbenzene	15.73	105	16528	0.4270 ug/L		92
79)	1,1,2,2-Tetrachloroethane	15.96	83	1999	0.3419 ug/L	#	83
83)	n-Propylbenzene	16.25	91	19410	0.4297 ug/L		95
84)	Bromobenzene	16.37	156	3319	0.3524 ug/L		75
85)	1,3,5-Trimethylbenzene	16.42	105	13794	0.4127 ug/L		95
86)	2-Chlorotoluene	16.51	91	13678	0.4729 ug/L		90
87)	4-Chlorotoluene	16.56	91	12572	0.4219 ug/L		98
88)	a-Methylstyrene	16.84	118	5739	0.3485 ug/L		96
89)	tert-Butylbenzene	16.90	134	2254	0.3403 ug/L		55
90)	1,2,4-Trimethylbenzene	16.95	105	13859	0.4051 ug/L		90
91)	sec-Butylbenzene	17.17	105	17612	0.4382 ug/L		90
92)	p-Isopropyltoluene	17.32	119	13616	0.3968 ug/L		90
93)	1,3-Dichlorobenzene	17.52	146	6961	0.3724 ug/L		86
94)	1,4-Dichlorobenzene	17.64	146	7149	0.3832 ug/L	#	1
95)	n-Butylbenzene	17.85	91	15223	0.4664 ug/L		91
96)	1,2-Dichlorobenzene	18.14	146	6129	0.3742 ug/L		92
98)	1,2,4-Trichlorobenzene	20.28	180	5035	0.3969 ug/L		97
99)	Hexachlorobutadiene	20.44	225	2420	0.4122 ug/L		83
100)	Naphthalene	20.65	128	7474	0.3874 ug/L	#	84
101)	1,2,3-Trichlorobenzene	20.96	180	4378	0.4056 ug/L	#	78

(#) = qualifier out of range (m) = manual integration
8M412221.D 8260WT.M Fri May 13 14:20:39 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412221.D Vial: 4
Acq On : 13 May 2016 13:58 Operator: TMB
Sample : WG568561-03 0.4ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:20 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412221.D Vial: 4
Acq On : 13 May 2016 13:58 Operator: TMB
Sample : WG568561-03 0.4ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:03:01 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	0.4000	0.4078	-2.0	100	0.00
3	P	Chloromethane	-1.0000	0.5012	0.0	100	0.00
4	C	Vinyl Chloride	0.4000	0.3775	5.6	100	0.00
5	T	1,3-Butadiene	-1.0000	0.3178	0.0	100	0.02
6	T	Bromomethane	-1.0000	1.0253	0.0	100	0.00
7	T	Chloroethane	0.4000	0.4335	-8.4	100	0.00
8	T	Trichlorofluoromethane	0.4000	0.4029	-0.7	100	-0.01
9	T	Diethyl ether	-1.0000	0.0000	0.0	0	-5.90#
10	T	Isoprene	0.4000	0.4079	-2.0	100	0.00
11	T	Acrolein	-1.0000	0.0000	0.0	0	-6.13#
12	T	1,1,2-Trichloro-1,2,2-Trifl	-1.0000	0.3894	0.0	100	0.02
13	T	Acetone	-1.0000	0.0000	0.0	0	-6.25#
14	C	1,1-Dichloroethene	0.4000	0.4192	-4.8	100	0.00
15	T	Tert-Butyl Alcohol	-1.0000	0.0000	0.0	0	-6.58#
16	T	Dimethyl Sulfide	0.4000	0.3687	7.8	100	0.00
17	T	Iodomethane	-1.0000	0.0000	0.0	0	-6.98#
18	T	Methyl acetate	-1.0000	0.2179	0.0	100	0.00
19	T	Methylene Chloride	-1.0000	0.4591	0.0	100	0.00
20	T	Carbon Disulfide	-1.0000	0.4587	0.0	100	0.00
21	T	Acrylonitrile	-1.0000	0.0000	0.0	0	-7.43#
22	T	Methyl Tert Butyl Ether	-1.0000	0.3688	0.0	100	0.00
23	T	trans-1,2-Dichloroethene	0.4000	0.4193	-4.8	100	0.00
24	T	n-Hexane	-1.0000	0.3519	0.0	100	0.00
25	T	Diisopropyl ether	-1.0000	0.0000	0.0	0	-8.13#
26	T	Vinyl Acetate	-1.0000	0.3241	0.0	100	0.02
27	P	1,1-Dichloroethane	0.4000	0.3826	4.4	100	0.00
28	T	Ethyl-Tert-Butyl ether	-1.0000	0.0000	0.0	0	-8.71#
29	T	2-Butanone	-1.0000	0.0000	0.0	0	-8.88#
30	T	Propionitrile	-1.0000	0.0000	0.0	0	-8.98#
31	T	2,2-Dichloropropane	0.4000	0.3874	3.1	100	0.00
32	T	cis-1,2-Dichloroethene	0.4000	0.4172	-4.3	100	-0.01
33	C	Chloroform	0.4000	0.4164	-4.1	100	0.00
34		1-Bromopropane	-1.0000	0.1776	0.0	100	-0.01
35	T	Bromochloromethane	0.4000	0.3550	11.3	100	0.00
36	T	Tetrahydrofuran	-1.0000	0.5125	0.0	100	-0.01
37	S	Dibromofluoromethane	-1.0000	0.0000	0.0	0	-9.67#
38	T	1,1,1-Trichloroethane	0.4000	0.3803	4.9	100	0.00
39	T	Cyclohexane	-1.0000	0.4058	0.0	100	0.00
40	T	1,1-Dichloropropene	-1.0000	0.4093	0.0	100	0.00
41	T	Tert-Amyl-Methyl ether	-1.0000	0.0000	0.0	0	-10.21#
42	T	Carbon Tetrachloride	0.4000	0.3867	3.3	100	0.00
43	S	1,2-Dichloroethane-d4	-1.0000	0.0000	0.0	0	-10.31#
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	0.4000	0.3955	1.1	100	0.00
46	T	Benzene	0.4000	0.4393	-9.8	100	0.00
47	T	Trichloroethene	0.4000	0.4169	-4.2	100	0.00
48	T	Methylcyclohexane	-1.0000	0.3768	0.0	100	-0.01
49	C	1,2-Dichloropropane	0.4000	0.4177	-4.4	100	0.00
50	T	Bromodichloromethane	0.4000	0.3676	8.1	100	0.00
51	T	1,4-Dioxane	-1.0000	0.0000	0.0	0	-11.71#
52	T	Dibromomethane	0.4000	0.4162	-4.1	100	0.00
53	T	2-Chloroethyl Vinyl Ether	-1.0000	0.1401	0.0	100	0.00
54	T	4-Methyl-2-Pentanone	-1.0000	0.0000	0.0	0	-12.05#

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412221.D Vial: 4
Acq On : 13 May 2016 13:58 Operator: TMB
Sample : WG568561-03 0.4ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:03:01 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	0.4000	0.3821	4.5	100	0.00
56	T	Dimethyl Disulfide	-1.0000	0.2746	0.0	100	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	-1.0000	0.0000	0.0	0	-12.68#
59	C	Toluene	0.4000	0.4259	-6.5	100	0.00
60	T	Ethyl Methacrylate	-1.0000	0.3102	0.0	100	0.02
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	-1.0000	0.3184	0.0	100	0.00
63	T	1,1,2-Trichloroethane	0.4000	0.4259	-6.5	100	0.00
64	T	2-Hexanone	-1.0000	0.0000	0.0	0	-13.12#
65	T	1,3-Dichloropropane	0.4000	0.3253	18.7	100	0.00
66	T	Tetrachloroethene	0.4000	0.4267	-6.7	100	0.00
67	T	Dibromochloromethane	0.4000	0.3904	2.4	100	0.00
68	T	1,2-Dibromoethane	0.4000	0.3814	4.6	100	0.00
69	T	1-Chlorohexane	0.4000	0.4017	-0.4	100	0.00
70	P	Chlorobenzene	0.4000	0.3933	1.7	100	0.00
71	T	1,1,1,2-Tetrachloroethane	0.4000	0.3872	3.2	100	0.00
72	C	Ethylbenzene	0.4000	0.4188	-4.7	100	0.00
73	T	m-,p-Xylene	0.8000	0.8859	-10.7	100	0.00
74	T	o-Xylene	-1.0000	0.4110	0.0	100	0.00
75	T	Styrene	0.4000	0.3901	2.5	100	0.00
76	P	Bromoform	-1.0000	0.2533	0.0	100	0.00
77	T	Isopropylbenzene	0.4000	0.4249	-6.2	100	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
79	P	1,1,2,2-Tetrachloroethane	0.4000	0.3398	15.1	100	0.00
80	S	p-Bromofluorobenzene	-1.0000	0.0283	0.0	100	0.00
81	T	1,2,3-Trichloropropane	-1.0000	0.0000	0.0	0	-16.14#
82	T	trans-1,4-Dichloro-2-Butene	-1.0000	0.0000	0.0	0	-16.19#
83	T	n-Propylbenzene	0.4000	0.4245	-6.1	100	0.00
84	T	Bromobenzene	0.4000	0.3529	11.8	100	0.00
85	T	1,3,5-Trimethylbenzene	0.4000	0.4141	-3.5	100	0.00
86	T	2-Chlorotoluene	0.4000	0.4284	-7.1	100	0.00
87	T	4-Chlorotoluene	0.4000	0.4359	-9.0	100	0.00
88	T	a-Methylstyrene	-1.0000	0.3631	0.0	100	0.00
89	T	tert-Butylbenzene	-1.0000	0.3459	0.0	100	0.00
90	T	1,2,4-Trimethylbenzene	0.4000	0.4046	-1.2	100	0.00
91	T	sec-Butylbenzene	-1.0000	0.4494	0.0	100	0.00
92	T	p-Isopropyltoluene	-1.0000	0.4092	0.0	100	0.00
93	T	1,3-Dichlorobenzene	0.4000	0.3844	3.9	100	0.00
94	T	1,4-Dichlorobenzene	0.4000	0.3915	2.1	100	0.00
95	T	n-Butylbenzene	-1.0000	0.4628	0.0	100	0.00
96	T	1,2-Dichlorobenzene	0.4000	0.3805	4.9	100	0.00
97	T	1,2-Dibromo-3-Chloropropane	-1.0000	0.0000	0.0	0	-19.13#
98	T	1,2,4-Trichlorobenzene	0.4000	0.4182	-4.5	100	0.00
99	T	Hexachlorobutadiene	0.4000	0.4302	-7.5	100	0.00
100	T	Naphthalene	0.4000	0.4178	-4.5	100	0.00
101	T	1,2,3-Trichlorobenzene	0.4000	0.4310	-7.7	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412221.D 8260WT.M Sat May 14 18:03:56 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412222.D Vial: 5
Acq On : 13 May 2016 14:27 Operator: TMB
Sample : WG568561-04 1ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:49:45 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	795577	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	547129	25.00 ug/L	0.01
78) 1,4-Dichlorobenzene-d4	17.60	152	284579	25.00 ug/L	0.01

System Monitoring Compounds					
37) Dibromofluoromethane	9.65	111	3962	0.3966 ug/L	-0.02
Spiked Amount 25.000	Range 86 - 118		Recovery	= 1.60%#	
43) 1,2-Dichloroethane-d4	10.31	65	3835	0.3356 ug/L	0.01
Spiked Amount 25.000	Range 80 - 120		Recovery	= 1.36%#	
58) Toluene-d8	12.68	98	14023	0.4957 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery	= 2.00%#	
80) p-Bromofluorobenzene	16.08	95	6483	0.5565 ug/L	0.01
Spiked Amount 25.000	Range 86 - 115		Recovery	= 2.24%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	12402	0.7438 ug/L	97
3) Chloromethane	3.62	50	8415	0.5357 ug/L #	75
4) Vinyl Chloride	3.84	62	9412	0.8873 ug/L	89
5) 1,3-Butadiene	3.90	54	6974	0.6729 ug/L	81
6) Bromomethane	4.73	94	4114	1.3123 ug/L	92
7) Chloroethane	4.90	64	4983	0.7178 ug/L #	80
8) Trichlorofluoromethane	5.38	101	17133	0.8136 ug/L #	96
9) Diethyl ether	5.90	59	20707	3.2140 ug/L	84
10) Isoprene	5.94	67	11176	1.0373 ug/L	76
11) Acrolein	6.14	56	246	0.4300 ug/L #	15
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	8135	0.8438 ug/L	96
13) Acetone	6.24	43	413	0.2115 ug/L #	48
14) 1,1-Dichloroethene	6.46	61	12947	0.6919 ug/L	91
15) Tert-Butyl Alcohol	6.59	59	3316	7.3747 ug/L #	55
16) Dimethyl Sulfide	6.72	62	4505	0.6515 ug/L	79
18) Methyl acetate	7.00	43	2119	0.3630 ug/L #	66
19) Methylene Chloride	7.24	84	8723	0.9837 ug/L #	43
20) Carbon Disulfide	7.29	76	28246	1.0441 ug/L	100
21) Acrylonitrile	7.43	53	2164	0.8620 ug/L	76
22) Methyl Tert Butyl Ether	7.47	73	16998	0.8678 ug/L	84
23) trans-1,2-Dichloroethene	7.70	61	12571	0.7427 ug/L	100
24) n-Hexane	7.78	57	9382	0.6109 ug/L	82
25) Diisopropyl ether	8.12	45	95400	2.1939 ug/L #	91
26) Vinyl Acetate	8.30	43	6750	0.3867 ug/L #	89
27) 1,1-Dichloroethane	8.32	63	14721	0.7271 ug/L	98
28) Ethyl-Tert-Butyl ether	8.70	59	97972	3.2102 ug/L	88
29) 2-Butanone	8.88	43	380	0.1215 ug/L #	57
30) Propionitrile	8.99	54	1593	2.0138 ug/L #	56
31) 2,2-Dichloropropane	9.09	77	14691	0.8236 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	9482	0.9362 ug/L	72
33) Chloroform	9.37	83	18494	0.9845 ug/L	94
34) 1-Bromopropane	9.50	122	1092	0.6844 ug/L	80
35) Bromochloromethane	9.61	130	5234	0.9877 ug/L #	55
36) Tetrahydrofuran	9.63	42	5415	2.6185 ug/L #	68
38) 1,1,1-Trichloroethane	9.91	97	17074	0.9016 ug/L	92
39) Cyclohexane	9.94	56	11865	0.5740 ug/L #	67
40) 1,1-Dichloropropene	10.11	75	13172	0.9663 ug/L	87
41) Tert-Amyl-Methyl ether	10.21	73	90255	4.6173 ug/L #	88
42) Carbon Tetrachloride	10.25	117	14515	0.8146 ug/L	95
45) 1,2-Dichloroethane	10.42	62	10675	0.6766 ug/L	85

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412222.D Vial: 5
Acq On : 13 May 2016 14:27 Operator: TMB
Sample : WG568561-04 1ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:49:45 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
46)	Benzene	10.47	78	34504	0.9938 ug/L		84
47)	Trichloroethene	11.21	130	8752	0.8735 ug/L		93
48)	Methylcyclohexane	11.30	83	13513	0.9646 ug/L		79
49)	1,2-Dichloropropane	11.43	63	7057	0.6982 ug/L	#	67
50)	Bromodichloromethane	11.73	83	12147	0.8638 ug/L		96
52)	Dibromomethane	11.81	93	4130	0.8681 ug/L		95
53)	2-Chloroethyl Vinyl Ether	12.03	63	2694	0.6366 ug/L		86
54)	4-Methyl-2-Pentanone	12.06	58	848	0.3671 ug/L	#	41
55)	cis-1,3-Dichloropropene	12.36	75	12521	0.8847 ug/L		97
56)	Dimethyl Disulfide	12.63	79	5799	0.7821 ug/L		96
59)	Toluene	12.79	91	37114	1.1328 ug/L		95
60)	Ethyl Methacrylate	12.88	69	5490	0.8930 ug/L		69
62)	trans-1,3-Dichloropropene	12.96	75	10328	0.9035 ug/L		92
63)	1,1,2-Trichloroethane	13.18	97	5274	1.0144 ug/L		93
64)	2-Hexanone	13.10	58	397	0.2127 ug/L	#	1
65)	1,3-Dichloropropane	13.49	76	9673	1.0032 ug/L		73
66)	Tetrachloroethene	13.61	164	7373	0.9937 ug/L		95
67)	Dibromochloromethane	13.87	129	7358	0.8977 ug/L		95
68)	1,2-Dibromoethane	14.13	107	5123	0.9454 ug/L		95
69)	1-Chlorohexane	14.21	91	12790	1.1461 ug/L		79
70)	Chlorobenzene	14.62	112	25662	1.1281 ug/L		89
71)	1,1,1,2-Tetrachloroethane	14.66	131	8722	0.9838 ug/L		94
72)	Ethylbenzene	14.65	106	13009	1.0078 ug/L		87
73)	m-,p-Xylene	14.75	106	30465	2.0486 ug/L		84
74)	o-Xylene	15.31	106	15133	1.0101 ug/L		84
75)	Styrene	15.35	104	23191	0.9707 ug/L		91
76)	Bromoform	15.84	173	3875	0.8240 ug/L		96
77)	Isopropylbenzene	15.74	105	41382	1.0895 ug/L		95
79)	1,1,2,2-Tetrachloroethane	15.95	83	5604	0.9873 ug/L	#	93
81)	1,2,3-Trichloropropane	16.15	110	1355	0.7428 ug/L	#	1
82)	trans-1,4-Dichloro-2-Buten	16.19	53	962	0.3479 ug/L	#	1
83)	n-Propylbenzene	16.25	91	48249	1.1002 ug/L		99
84)	Bromobenzene	16.37	156	9598	1.0497 ug/L		95
85)	1,3,5-Trimethylbenzene	16.42	105	33355	1.0278 ug/L		96
86)	2-Chlorotoluene	16.51	91	33408	1.1895 ug/L		94
87)	4-Chlorotoluene	16.56	91	29548	1.0213 ug/L		95
88)	a-Methylstyrene	16.84	118	14414	0.9014 ug/L		91
89)	tert-Butylbenzene	16.89	134	6686	1.0398 ug/L		86
90)	1,2,4-Trimethylbenzene	16.95	105	36066	1.0857 ug/L		96
91)	sec-Butylbenzene	17.16	105	42128	1.0796 ug/L		96
92)	p-Isopropyltoluene	17.32	119	35424	1.0633 ug/L		94
93)	1,3-Dichlorobenzene	17.52	146	18409	1.0144 ug/L		93
94)	1,4-Dichlorobenzene	17.64	146	19031	1.0506 ug/L	#	62
95)	n-Butylbenzene	17.85	91	35772	1.1288 ug/L		96
96)	1,2-Dichlorobenzene	18.14	146	16104	1.0126 ug/L		92
97)	1,2-Dibromo-3-Chloropropan	19.14	75	1262	1.0152 ug/L	#	53
98)	1,2,4-Trichlorobenzene	20.28	180	11494	0.9331 ug/L		98
99)	Hexachlorobutadiene	20.43	225	5421	0.9510 ug/L		98
100)	Naphthalene	20.65	128	16581	0.8853 ug/L		100
101)	1,2,3-Trichlorobenzene	20.96	180	10306	0.9835 ug/L		94

(#) = qualifier out of range (m) = manual integration
8M412222.D 8260WT.M Fri May 13 14:49:48 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412222.D Vial: 5
Acq On : 13 May 2016 14:27 Operator: TMB
Sample : WG568561-04 1ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 14:49 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412222.D Vial: 5
Acq On : 13 May 2016 14:27 Operator: TMB
Sample : WG568561-04 1ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:05:31 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	1.0000	0.9298	7.0	100	0.00
3	P	Chloromethane	1.0000	1.0808	-8.1	100	0.00
4	C	Vinyl Chloride	1.0000	1.0189	-1.9	100	0.00
5	T	1,3-Butadiene	-1.0000	0.9002	0.0	100	0.02
6	T	Bromomethane	1.0000	1.2082	-20.8	100	0.00
7	T	Chloroethane	1.0000	0.9803	2.0	100	0.00
8	T	Trichlorofluoromethane	1.0000	0.9968	0.3	100	0.00
9	T	Diethyl ether	5.0000	5.0462	-0.9	100	0.00
10	T	Isoprene	1.0000	1.0707	-7.1	100	0.00
11	T	Acrolein	-1.0000	0.6696	0.0	100	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	1.0000	1.0069	-0.7	100	0.00
13	T	Acetone	-1.0000	0.4614	0.0	100	0.00
14	C	1,1-Dichloroethene	1.0000	1.0064	-0.6	100	0.00
15	T	Tert-Butyl Alcohol	-1.0000	11.5604	0.0	0	0.02
16	T	Dimethyl Sulfide	1.0000	0.9794	2.1	100	-0.01
17	T	Iodomethane	-1.0000	0.0000	0.0	0	-6.98#
18	T	Methyl acetate	-1.0000	0.8758	0.0	100	0.00
19	T	Methylene Chloride	1.0000	1.0655	-6.5	100	0.00
20	T	Carbon Disulfide	1.0000	1.0492	-4.9	100	0.00
21	T	Acrylonitrile	2.5000	1.8633	25.5#	100	0.00
22	T	Methyl Tert Butyl Ether	1.0000	0.9593	4.1	100	0.00
23	T	trans-1,2-Dichloroethene	1.0000	1.0636	-6.4	100	0.00
24	T	n-Hexane	-1.0000	1.0690	0.0	100	0.00
25	T	Diisopropyl ether	5.0000	5.1784	-3.6	100	0.00
26	T	Vinyl Acetate	-1.0000	0.8002	0.0	100	0.00
27	P	1,1-Dichloroethane	1.0000	1.0029	-0.3	100	0.00
28	T	Ethyl-Tert-Butyl ether	5.0000	4.9635	0.7	100	0.00
29	T	2-Butanone	-1.0000	0.2889	0.0	100	0.00
30	T	Propionitrile	5.0000	4.1566	16.9	100	0.00
31	T	2,2-Dichloropropane	1.0000	0.9498	5.0	100	0.00
32	T	cis-1,2-Dichloroethene	1.0000	1.0383	-3.8	100	0.00
33	C	Chloroform	1.0000	1.0671	-6.7	100	0.00
34		1-Bromopropane	1.0000	0.8348	16.5	100	-0.01
35	T	Bromochloromethane	1.0000	1.1177	-11.8	100	0.00
36	T	Tetrahydrofuran	5.0000	6.2516	-25.0#	100	0.00
37	S	Dibromofluoromethane	0.5000	0.4635	7.3	100	-0.01
38	T	1,1,1-Trichloroethane	1.0000	1.0372	-3.7	100	0.00
39	T	Cyclohexane	1.0000	1.0411	-4.1	100	0.00
40	T	1,1-Dichloropropene	1.0000	1.0498	-5.0	100	0.00
41	T	Tert-Amyl-Methyl ether	5.0000	5.0344	-0.7	100	0.00
42	T	Carbon Tetrachloride	1.0000	0.9930	0.7	100	0.00
43	S	1,2-Dichloroethane-d4	0.5000	0.4184	16.3	100	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	1.0000	0.9570	4.3	100	0.00
46	T	Benzene	1.0000	1.0558	-5.6	100	0.00
47	T	Trichloroethene	1.0000	1.0041	-0.4	100	0.00
48	T	Methylcyclohexane	-1.0000	1.0205	0.0	100	0.00
49	C	1,2-Dichloropropane	1.0000	0.9700	3.0	100	0.00
50	T	Bromodichloromethane	1.0000	0.9500	5.0	100	0.00
51	T	1,4-Dioxane	-1.0000	0.0000	0.0	0	-11.71#
52	T	Dibromomethane	1.0000	0.9590	4.1	100	0.00
53	T	2-Chloroethyl Vinyl Ether	-1.0000	0.8616	0.0	100	0.00
54	T	4-Methyl-2-Pentanone	-1.0000	0.6639	0.0	100	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412222.D Vial: 5
Acq On : 13 May 2016 14:27 Operator: TMB
Sample : WG568561-04 1ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:05:31 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	1.0000	0.9621	3.8	100	0.00
56	T	Dimethyl Disulfide	-1.0000	0.8348	0.0	100	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	0.5000	0.4897	2.1	100	0.00
59	C	Toluene	1.0000	1.0927	-9.3	100	0.00
60	T	Ethyl Methacrylate	1.0000	0.8416	15.8	100	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	1.0000	0.9073	9.3	100	0.00
63	T	1,1,2-Trichloroethane	1.0000	0.9840	1.6	100	0.00
64	T	2-Hexanone	-1.0000	0.3525	0.0	100	-0.01
65	T	1,3-Dichloropropane	1.0000	1.0137	-1.4	100	0.00
66	T	Tetrachloroethene	1.0000	1.0132	-1.3	100	0.00
67	T	Dibromochloromethane	1.0000	0.9402	6.0	100	0.00
68	T	1,2-Dibromoethane	1.0000	0.9590	4.1	100	0.00
69	T	1-Chlorohexane	1.0000	1.1031	-10.3	100	0.00
70	P	Chlorobenzene	1.0000	1.1204	-12.0	100	0.00
71	T	1,1,1,2-Tetrachloroethane	1.0000	0.9829	1.7	100	0.00
72	C	Ethylbenzene	1.0000	1.0381	-3.8	100	0.00
73	T	m-,p-Xylene	2.0000	2.0726	-3.6	100	0.00
74	T	o-Xylene	1.0000	1.0530	-5.3	100	0.00
75	T	Styrene	1.0000	0.9839	1.6	100	0.00
76	P	Bromoform	1.0000	0.8348	16.5	100	0.00
77	T	Isopropylbenzene	1.0000	1.0840	-8.4	100	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
79	P	1,1,2,2-Tetrachloroethane	1.0000	0.9810	1.9	100	0.00
80	S	p-Bromofluorobenzene	-1.0000	0.5502	0.0	100	0.00
81	T	1,2,3-Trichloropropane	1.0000	0.7703	23.0	100	0.02
82	T	trans-1,4-Dichloro-2-Butene	1.0000	1.1443	-14.4	100	0.00
83	T	n-Propylbenzene	1.0000	1.0868	-8.7	100	0.00
84	T	Bromobenzene	1.0000	1.0510	-5.1	100	0.00
85	T	1,3,5-Trimethylbenzene	1.0000	1.0314	-3.1	100	0.00
86	T	2-Chlorotoluene	1.0000	1.0778	-7.8	100	0.00
87	T	4-Chlorotoluene	1.0000	1.0553	-5.5	100	0.00
88	T	a-Methylstyrene	-1.0000	0.9394	0.0	100	0.00
89	T	tert-Butylbenzene	1.0000	1.0567	-5.7	100	0.00
90	T	1,2,4-Trimethylbenzene	1.0000	1.0845	-8.5	100	0.00
91	T	sec-Butylbenzene	1.0000	1.1073	-10.7	100	0.00
92	T	p-Isopropyltoluene	1.0000	1.0966	-9.7	100	0.00
93	T	1,3-Dichlorobenzene	1.0000	1.0469	-4.7	100	0.00
94	T	1,4-Dichlorobenzene	1.0000	1.0735	-7.3	100	0.00
95	T	n-Butylbenzene	1.0000	1.1201	-12.0	100	0.00
96	T	1,2-Dichlorobenzene	1.0000	1.0297	-3.0	100	0.00
97	T	1,2-Dibromo-3-Chloropropane	-1.0000	1.1006	0.0	100	0.00
98	T	1,2,4-Trichlorobenzene	1.0000	0.9832	1.7	100	0.00
99	T	Hexachlorobutadiene	1.0000	0.9925	0.7	100	0.00
100	T	Naphthalene	1.0000	0.9547	4.5	100	0.00
101	T	1,2,3-Trichlorobenzene	1.0000	1.0450	-4.5	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412223.D Vial: 6
Acq On : 13 May 2016 14:56 Operator: TMB
Sample : WG568561-05 2ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:18:34 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	10.71	96	782460	25.00	ug/L	0.00
57) Chlorobenzene-d5	14.57	117	543252	25.00	ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	286450	25.00	ug/L	0.01

System Monitoring Compounds						
37) Dibromofluoromethane	9.66	111	8161	0.8307	ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery =	3.32%#		
43) 1,2-Dichloroethane-d4	10.30	65	9681	0.8613	ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery =	3.44%#		
58) Toluene-d8	12.68	98	28135	1.0016	ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery =	4.00%#		
80) p-Bromofluorobenzene	16.08	95	12079	1.0302	ug/L	0.01
Spiked Amount 25.000	Range 86 - 115		Recovery =	4.12%#		

Target Compounds						Qvalue
2) Dichlorodifluoromethane	3.16	85	24870	1.5166	ug/L	98
3) Chloromethane	3.62	50	16106	1.0425	ug/L	93
4) Vinyl Chloride	3.85	62	18562	1.7792	ug/L	94
5) 1,3-Butadiene	3.90	54	13617	1.3360	ug/L	92
6) Bromomethane	4.73	94	8329	1.8804	ug/L	94
7) Chloroethane	4.90	64	9562	1.4006	ug/L	86
8) Trichlorofluoromethane	5.39	101	34970	1.6885	ug/L	97
9) Diethyl ether	5.90	59	107986	17.0420	ug/L	79
10) Isoprene	5.93	67	20936	1.9757	ug/L	73
11) Acrolein	6.14	56	4864	8.6444	ug/L	95
12) 1,1,2-Trichloro-1,2,2-Trif	6.17	101	16832	1.7751	ug/L	99
13) Acetone	6.23	43	1519	0.7908	ug/L #	48
14) 1,1-Dichloroethene	6.47	61	25394	1.3798	ug/L	89
15) Tert-Butyl Alcohol	6.58	59	14891	33.6725	ug/L	85
16) Dimethyl Sulfide	6.73	62	9489	1.3953	ug/L	79
17) Iodomethane	6.98	142	1211	1.3472	ug/L #	60
18) Methyl acetate	7.01	43	5064	0.8821	ug/L #	79
19) Methylene Chloride	7.25	84	16579	1.9010	ug/L	52
20) Carbon Disulfide	7.29	76	57693	2.1683	ug/L	98
21) Acrylonitrile	7.43	53	13848	5.6089	ug/L	93
22) Methyl Tert Butyl Ether	7.47	73	33991	1.7644	ug/L	84
23) trans-1,2-Dichloroethene	7.70	61	22786	1.3688	ug/L	88
24) n-Hexane	7.79	57	18282	1.2105	ug/L	84
25) Diisopropyl ether	8.12	45	475271	11.1128	ug/L #	92
26) Vinyl Acetate	8.29	43	15998	0.9318	ug/L	93
27) 1,1-Dichloroethane	8.32	63	30419	1.5277	ug/L	97
28) Ethyl-Tert-Butyl ether	8.70	59	505010	16.8250	ug/L	89
29) 2-Butanone	8.90	43	2250	0.7313	ug/L #	57
30) Propionitrile	8.99	54	9373	12.0479	ug/L	86
31) 2,2-Dichloropropane	9.10	77	32521	1.8538	ug/L	99
32) cis-1,2-Dichloroethene	9.17	96	18238	1.8309	ug/L	91
33) Chloroform	9.37	83	33543	1.8156	ug/L	99
34) 1-Bromopropane	9.52	122	2513	1.6014	ug/L	88
35) Bromochloromethane	9.60	130	9189	1.7632	ug/L	67
36) Tetrahydrofuran	9.63	42	22074	10.8529	ug/L #	66
38) 1,1,1-Trichloroethane	9.90	97	32268	1.7324	ug/L	96
39) Cyclohexane	9.93	56	23097	1.1362	ug/L #	65
40) 1,1-Dichloropropene	10.10	75	24988	1.8639	ug/L	99
41) Tert-Amyl-Methyl ether	10.21	73	459535	23.9029	ug/L	87
42) Carbon Tetrachloride	10.24	117	28935	1.6512	ug/L	99

(#) = qualifier out of range (m) = manual integration
8M412223.D 8260WT.M Fri May 13 15:18:36 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412223.D Vial: 6
Acq On : 13 May 2016 14:56 Operator: TMB
Sample : WG568561-05 2ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:18:34 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit	Qvalue
45)	1,2-Dichloroethane	10.42	62	22142	1.4270 ug/L #	83
46)	Benzene	10.46	78	67028	1.9629 ug/L	83
47)	Trichloroethene	11.21	130	17287	1.7543 ug/L	96
48)	Methylcyclohexane	11.30	83	26050	1.8907 ug/L	84
49)	1,2-Dichloropropane	11.42	63	14582	1.4670 ug/L	77
50)	Bromodichloromethane	11.73	83	25794	1.8651 ug/L	93
51)	1,4-Dioxane	11.73	88	1282	32.3515 ug/L	74
52)	Dibromomethane	11.81	93	8575	1.8325 ug/L	97
53)	2-Chloroethyl Vinyl Ether	12.02	63	6234	1.4979 ug/L	100
54)	4-Methyl-2-Pentanone	12.06	58	2139	0.9416 ug/L	84
55)	cis-1,3-Dichloropropene	12.37	75	24153	1.7353 ug/L	98
56)	Dimethyl Disulfide	12.62	79	12288	1.6851 ug/L	99
59)	Toluene	12.79	91	72944	2.2423 ug/L	95
60)	Ethyl Methacrylate	12.88	69	12716	2.0832 ug/L	71
62)	trans-1,3-Dichloropropene	12.96	75	22163	1.9527 ug/L	100
63)	1,1,2-Trichloroethane	13.18	97	10537	2.0412 ug/L	99
64)	2-Hexanone	13.13	58	2200	1.1873 ug/L #	65
65)	1,3-Dichloropropane	13.49	76	20212	2.1112 ug/L	72
66)	Tetrachloroethene	13.60	164	14595	1.9811 ug/L	90
67)	Dibromochloromethane	13.87	129	14698	1.8061 ug/L	98
68)	1,2-Dibromoethane	14.12	107	10152	1.8869 ug/L	95
69)	1-Chlorohexane	14.21	91	21960	1.9818 ug/L	89
70)	Chlorobenzene	14.62	112	47376	2.0974 ug/L	81
71)	1,1,1,2-Tetrachloroethane	14.66	131	17526	1.9910 ug/L	97
72)	Ethylbenzene	14.66	106	25628	1.9996 ug/L	92
73)	m-,p-Xylene	14.74	106	60087	4.0693 ug/L	84
74)	o-Xylene	15.31	106	28289	1.9017 ug/L	79
75)	Styrene	15.35	104	47644	2.0086 ug/L	92
76)	Bromoform	15.84	173	9145	1.9585 ug/L	95
77)	Isopropylbenzene	15.73	105	81301	2.1557 ug/L	97
79)	1,1,2,2-Tetrachloroethane	15.95	83	11538	2.0194 ug/L	95
81)	1,2,3-Trichloropropane	16.13	110	3630	1.9770 ug/L	94
82)	trans-1,4-Dichloro-2-Buten	16.19	53	2699	0.9698 ug/L #	36
83)	n-Propylbenzene	16.24	91	97233	2.2026 ug/L	96
84)	Bromobenzene	16.37	156	19174	2.0834 ug/L	92
85)	1,3,5-Trimethylbenzene	16.42	105	67395	2.0631 ug/L	97
86)	2-Chlorotoluene	16.51	91	67110	2.3739 ug/L	94
87)	4-Chlorotoluene	16.56	91	59154	2.0313 ug/L	92
88)	a-Methylstyrene	16.84	118	29415	1.8275 ug/L	86
89)	tert-Butylbenzene	16.89	134	12716	1.9647 ug/L	67
90)	1,2,4-Trimethylbenzene	16.95	105	71669	2.1435 ug/L	94
91)	sec-Butylbenzene	17.17	105	81038	2.0632 ug/L	98
92)	p-Isopropyltoluene	17.32	119	70170	2.0924 ug/L	92
93)	1,3-Dichlorobenzene	17.52	146	37289	2.0413 ug/L	93
94)	1,4-Dichlorobenzene	17.64	146	36528	2.0034 ug/L #	73
95)	n-Butylbenzene	17.85	91	67823	2.1262 ug/L	96
96)	1,2-Dichlorobenzene	18.14	146	32322	2.0190 ug/L	92
97)	1,2-Dibromo-3-Chloropropan	19.13	75	1955	1.5624 ug/L	94
98)	1,2,4-Trichlorobenzene	20.28	180	23558	1.9000 ug/L	98
99)	Hexachlorobutadiene	20.43	225	11066	1.9286 ug/L	99
100)	Naphthalene	20.65	128	36066	1.9130 ug/L	99
101)	1,2,3-Trichlorobenzene	20.96	180	19765	1.8738 ug/L	95

(#) = qualifier out of range (m) = manual integration
8M412223.D 8260WT.M Fri May 13 15:18:37 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412223.D Vial: 6
Acq On : 13 May 2016 14:56 Operator: TMB
Sample : WG568561-05 2ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:18 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration



8M412223.D 8260WT.M Fri May 13 15:18:37 2016 Page 3

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412223.D Vial: 6
Acq On : 13 May 2016 14:56 Operator: TMB
Sample : WG568561-05 2ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	2.0000	1.8958	5.2	100	0.00
3	P	Chloromethane	2.0000	2.1033	-5.2	100	0.00
4	C	Vinyl Chloride	2.0000	2.0431	-2.2	100	0.00
5	T	1,3-Butadiene	2.0000	1.7872	10.6	100	0.02
6	T	Bromomethane	2.0000	2.0126	-0.6	100	0.00
7	T	Chloroethane	2.0000	1.9127	4.4	100	0.00
8	T	Trichlorofluoromethane	2.0000	2.0686	-3.4	100	0.00
9	T	Diethyl ether	25.0000	26.7567	-7.0	100	0.00
10	T	Isoprene	2.0000	2.0394	-2.0	100	0.00
11	T	Acrolein	12.5000	13.4606	-7.7	100	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	2.0000	2.1182	-5.9	100	0.02
13	T	Acetone	-1.0000	1.7255	0.0	100	-0.01
14	C	1,1-Dichloroethene	2.0000	2.0070	-0.4	100	0.00
15	T	Tert-Butyl Alcohol	50.0000	52.7840	-5.6	100	0.00
16	T	Dimethyl Sulfide	2.0000	2.0975	-4.9	100	0.00
17	T	Iodomethane	2.0000	3.0680	-53.4#	100	0.00
18	T	Methyl acetate	-1.0000	2.1281	0.0	100	0.00
19	T	Methylene Chloride	2.0000	2.0590	-3.0	100	0.00
20	T	Carbon Disulfide	2.0000	2.1789	-8.9	100	0.00
21	T	Acrylonitrile	12.5000	12.1237	3.0	100	0.00
22	T	Methyl Tert Butyl Ether	2.0000	1.9505	2.5	100	0.00
23	T	trans-1,2-Dichloroethene	2.0000	1.9603	2.0	100	0.00
24	T	n-Hexane	-1.0000	2.1181	0.0	100	0.00
25	T	Diisopropyl ether	25.0000	26.2305	-4.9	100	0.00
26	T	Vinyl Acetate	2.0000	1.9284	3.6	100	0.00
27	P	1,1-Dichloroethane	2.0000	2.1072	-5.4	100	0.00
28	T	Ethyl-Tert-Butyl ether	25.0000	26.0139	-4.1	100	0.00
29	T	2-Butanone	-1.0000	1.7396	0.0	100	0.02
30	T	Propionitrile	25.0000	24.8668	0.5	100	0.00
31	T	2,2-Dichloropropane	2.0000	2.1378	-6.9	100	0.00
32	T	cis-1,2-Dichloroethene	2.0000	2.0306	-1.5	100	0.00
33	C	Chloroform	2.0000	1.9679	1.6	100	0.00
34		1-Bromopropane	2.0000	1.9533	2.3	100	0.00
35	T	Bromochloromethane	2.0000	1.9951	0.2	100	0.00
36	T	Tetrahydrofuran	25.0000	25.9117	-3.6	100	0.00
37	S	Dibromofluoromethane	1.0000	0.9708	2.9	100	0.00
38	T	1,1,1-Trichloroethane	2.0000	1.9930	0.3	100	0.00
39	T	Cyclohexane	2.0000	2.0605	-3.0	100	0.00
40	T	1,1-Dichloropropene	2.0000	2.0248	-1.2	100	0.00
41	T	Tert-Amyl-Methyl ether	25.0000	26.0627	-4.3	100	0.00
42	T	Carbon Tetrachloride	2.0000	2.0126	-0.6	100	0.00
43	S	1,2-Dichloroethane-d4	1.0000	1.0740	-7.4	100	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	2.0000	2.0183	-0.9	100	0.00
46	T	Benzene	2.0000	2.0854	-4.3	100	0.00
47	T	Trichloroethene	2.0000	2.0166	-0.8	100	0.00
48	T	Methylcyclohexane	-1.0000	2.0003	0.0	100	0.00
49	C	1,2-Dichloropropane	2.0000	2.0380	-1.9	100	0.00
50	T	Bromodichloromethane	2.0000	2.0511	-2.6	100	0.00
51	T	1,4-Dioxane	-1.0000	35.6688	0.0	100	0.02
52	T	Dibromomethane	2.0000	2.0246	-1.2	100	0.00
53	T	2-Chloroethyl Vinyl Ether	2.0000	2.0271	-1.4	100	0.00
54	T	4-Methyl-2-Pentanone	-1.0000	1.7027	0.0	100	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412223.D Vial: 6
Acq On : 13 May 2016 14:56 Operator: TMB
Sample : WG568561-05 2ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	2.0000	1.8869	5.7	100	0.00
56	T	Dimethyl Disulfide	-1.0000	1.7986	0.0	100	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	1.0000	0.9896	1.0	100	0.00
59	C	Toluene	2.0000	2.1628	-8.1	100	0.00
60	T	Ethyl Methacrylate	2.0000	1.9633	1.8	100	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	2.0000	1.9609	2.0	100	0.00
63	T	1,1,2-Trichloroethane	2.0000	1.9800	1.0	100	0.00
64	T	2-Hexanone	-1.0000	1.9672	0.0	100	0.00
65	T	1,3-Dichloropropane	2.0000	2.1332	-6.7	100	0.00
66	T	Tetrachloroethene	2.0000	2.0198	-1.0	100	0.00
67	T	Dibromochloromethane	2.0000	1.8915	5.4	100	0.00
68	T	1,2-Dibromoethane	2.0000	1.9141	4.3	100	0.00
69	T	1-Chlorohexane	2.0000	1.9075	4.6	100	0.00
70	P	Chlorobenzene	2.0000	2.0831	-4.2	100	0.00
71	T	1,1,1,2-Tetrachloroethane	2.0000	1.9892	0.5	100	0.00
72	C	Ethylbenzene	2.0000	2.0596	-3.0	100	0.00
73	T	m-,p-Xylene	4.0000	4.1170	-2.9	100	0.00
74	T	o-Xylene	2.0000	1.9825	0.9	100	0.00
75	T	Styrene	2.0000	2.0358	-1.8	100	0.00
76	P	Bromoform	2.0000	1.9841	0.8	100	0.00
77	T	Isopropylbenzene	2.0000	2.1448	-7.2	100	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
79	P	1,1,2,2-Tetrachloroethane	2.0000	2.0066	-0.3	100	0.00
80	S	p-Bromofluorobenzene	1.0000	1.0184	-1.8	100	0.00
81	T	1,2,3-Trichloropropane	2.0000	2.0501	-2.5	100	0.00
82	T	trans-1,4-Dichloro-2-Butene	2.0000	2.0915	-4.6	100	0.00
83	T	n-Propylbenzene	2.0000	2.1757	-8.8	100	0.00
84	T	Bromobenzene	2.0000	2.0859	-4.3	100	0.00
85	T	1,3,5-Trimethylbenzene	2.0000	2.0703	-3.5	100	0.00
86	T	2-Chlorotoluene	2.0000	2.1509	-7.5	100	0.00
87	T	4-Chlorotoluene	2.0000	2.0989	-4.9	100	0.00
88	T	a-Methylstyrene	-1.0000	1.9045	0.0	100	0.00
89	T	tert-Butylbenzene	2.0000	1.9967	0.2	100	0.00
90	T	1,2,4-Trimethylbenzene	2.0000	2.1411	-7.1	100	0.00
91	T	sec-Butylbenzene	2.0000	2.1161	-5.8	100	0.00
92	T	p-Isopropyltoluene	2.0000	2.1579	-7.9	100	0.00
93	T	1,3-Dichlorobenzene	2.0000	2.1067	-5.3	100	0.00
94	T	1,4-Dichlorobenzene	2.0000	2.0470	-2.4	100	0.00
95	T	n-Butylbenzene	2.0000	2.1098	-5.5	100	0.00
96	T	1,2-Dichlorobenzene	2.0000	2.0532	-2.7	100	0.00
97	T	1,2-Dibromo-3-Chloropropane	2.0000	1.6938	15.3	100	0.00
98	T	1,2,4-Trichlorobenzene	2.0000	2.0021	-0.1	100	0.00
99	T	Hexachlorobutadiene	2.0000	2.0128	-0.6	100	0.00
100	T	Naphthalene	2.0000	2.0630	-3.2	100	0.00
101	T	1,2,3-Trichlorobenzene	2.0000	1.9910	0.4	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412224.D Vial: 7
Acq On : 13 May 2016 15:25 Operator: TMB
Sample : WG568561-06 5ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:47:41 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	768397	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	534372	25.00 ug/L	0.01
78) 1,4-Dichlorobenzene-d4	17.60	152	274234	25.00 ug/L	0.01

System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	19775	2.0496 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery =	8.20%#	
43) 1,2-Dichloroethane-d4	10.30	65	21200	1.9207 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery =	7.68%#	
58) Toluene-d8	12.68	98	65647	2.3758 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery =	9.52%#	
80) p-Bromofluorobenzene	16.08	95	26759	2.3838 ug/L	0.01
Spiked Amount 25.000	Range 86 - 115		Recovery =	9.52%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	55961	3.4750 ug/L	99
3) Chloromethane	3.62	50	34194	2.2538 ug/L	100
4) Vinyl Chloride	3.84	62	42506	4.1488 ug/L	99
5) 1,3-Butadiene	3.89	54	32006	3.1976 ug/L	93
6) Bromomethane	4.72	94	19582	3.4204 ug/L	94
7) Chloroethane	4.90	64	22000	3.2814 ug/L	84
8) Trichlorofluoromethane	5.37	101	77715	3.8212 ug/L	96
9) Diethyl ether	5.90	59	188516	30.2955 ug/L	79
10) Isoprene	5.94	67	44975	4.3218 ug/L	78
11) Acrolein	6.15	56	7712	13.9568 ug/L	92
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	36867	3.9591 ug/L	99
13) Acetone	6.24	43	3578	1.8967 ug/L	85
14) 1,1-Dichloroethene	6.47	61	59405	3.2870 ug/L	95
15) Tert-Butyl Alcohol	6.58	59	27833	64.0897 ug/L #	82
16) Dimethyl Sulfide	6.73	62	21102	3.1597 ug/L	77
17) Iodomethane	6.99	142	3667	1.8964 ug/L	95
18) Methyl acetate	7.00	43	11500	2.0398 ug/L #	86
19) Methylene Chloride	7.24	84	37533	4.3823 ug/L	52
20) Carbon Disulfide	7.29	76	123060	4.7097 ug/L	100
21) Acrylonitrile	7.43	53	25653	10.5804 ug/L	99
22) Methyl Tert Butyl Ether	7.47	73	80005	4.2289 ug/L	86
23) trans-1,2-Dichloroethene	7.70	61	54124	3.3108 ug/L	94
24) n-Hexane	7.78	57	40758	2.7480 ug/L	82
25) Diisopropyl ether	8.12	45	841400	20.0337 ug/L #	91
26) Vinyl Acetate	8.30	43	37284	2.2113 ug/L #	90
27) 1,1-Dichloroethane	8.32	63	66711	3.4116 ug/L	97
28) Ethyl-Tert-Butyl ether	8.70	59	907423	30.7852 ug/L	89
29) 2-Butanone	8.89	43	5359	1.7738 ug/L #	82
30) Propionitrile	8.99	54	17933	23.4726 ug/L	100
31) 2,2-Dichloropropane	9.10	77	73150	4.2462 ug/L	99
32) cis-1,2-Dichloroethene	9.17	96	41036	4.1950 ug/L	89
33) Chloroform	9.37	83	78227	4.3116 ug/L	99
34) 1-Bromopropane	9.52	122	6162	3.9986 ug/L	99
35) Bromochloromethane	9.60	130	20988	4.1009 ug/L #	62
36) Tetrahydrofuran	9.62	42	38098	19.0742 ug/L #	62
38) 1,1,1-Trichloroethane	9.91	97	76152	4.1633 ug/L	97
39) Cyclohexane	9.94	56	51439	2.5767 ug/L #	66
40) 1,1-Dichloropropene	10.11	75	56645	4.3027 ug/L	93
41) Tert-Amyl-Methyl ether	10.21	73	821180	43.4958 ug/L	87
42) Carbon Tetrachloride	10.25	117	67831	3.9416 ug/L	96

(#) = qualifier out of range (m) = manual integration
8M412224.D 8260WT.M Fri May 13 15:47:43 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412224.D Vial: 7
Acq On : 13 May 2016 15:25 Operator: TMB
Sample : WG568561-06 5ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:47:41 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
45)	1,2-Dichloroethane	10.43	62	51708	3.3934 ug/L		85
46)	Benzene	10.46	78	150204	4.4792 ug/L		86
47)	Trichloroethene	11.21	130	38500	3.9784 ug/L		94
48)	Methylcyclohexane	11.30	83	60897	4.5008 ug/L		81
49)	1,2-Dichloropropane	11.43	63	31773	3.2549 ug/L		77
50)	Bromodichloromethane	11.73	83	59127	4.3535 ug/L		94
51)	1,4-Dioxane	11.71	88	2719	69.8702 ug/L	#	62
52)	Dibromomethane	11.81	93	19711	4.2894 ug/L		96
53)	2-Chloroethyl Vinyl Ether	12.03	63	12946	3.1676 ug/L		98
54)	4-Methyl-2-Pentanone	12.05	58	5129	2.2991 ug/L	#	78
55)	cis-1,3-Dichloropropene	12.36	75	61169	4.4752 ug/L		94
56)	Dimethyl Disulfide	12.62	79	27966	3.9053 ug/L		98
59)	Toluene	12.78	91	161007	5.0316 ug/L		97
60)	Ethyl Methacrylate	12.88	69	28595	4.7624 ug/L		71
62)	trans-1,3-Dichloropropene	12.96	75	49791	4.4599 ug/L		97
63)	1,1,2-Trichloroethane	13.18	97	23678	4.6631 ug/L		99
64)	2-Hexanone	13.11	58	4499	2.4685 ug/L	#	97
65)	1,3-Dichloropropane	13.49	76	44838	4.7614 ug/L		73
66)	Tetrachloroethene	13.60	164	32715	4.5144 ug/L		97
67)	Dibromochloromethane	13.87	129	35009	4.3734 ug/L		98
68)	1,2-Dibromoethane	14.13	107	24247	4.5815 ug/L		99
69)	1-Chlorohexane	14.21	91	53599	4.9176 ug/L		84
70)	Chlorobenzene	14.62	112	107618	4.8437 ug/L		95
71)	1,1,1,2-Tetrachloroethane	14.65	131	38120	4.4025 ug/L		96
72)	Ethylbenzene	14.65	106	56905	4.5138 ug/L		91
73)	m-,p-Xylene	14.75	106	137390	9.4592 ug/L		86
74)	o-Xylene	15.31	106	65650	4.4866 ug/L		85
75)	Styrene	15.35	104	111433	4.7758 ug/L		91
76)	Bromoform	15.84	173	19346	4.2121 ug/L		97
77)	Isopropylbenzene	15.73	105	182523	4.9200 ug/L		96
79)	1,1,2,2-Tetrachloroethane	15.95	83	26655	4.8730 ug/L		99
81)	1,2,3-Trichloropropane	16.13	110	8840	5.0289 ug/L		81
82)	trans-1,4-Dichloro-2-Buten	16.19	53	6584	2.4712 ug/L	#	51
83)	n-Propylbenzene	16.24	91	222196	5.2576 ug/L		96
84)	Bromobenzene	16.37	156	42665	4.8423 ug/L		95
85)	1,3,5-Trimethylbenzene	16.43	105	155714	4.9790 ug/L		93
86)	2-Chlorotoluene	16.52	91	148764	5.4968 ug/L		94
87)	4-Chlorotoluene	16.56	91	135679	4.8666 ug/L		94
88)	a-Methylstyrene	16.84	118	68945	4.4743 ug/L		91
89)	tert-Butylbenzene	16.89	134	29069	4.6914 ug/L		77
90)	1,2,4-Trimethylbenzene	16.95	105	161016	5.0301 ug/L		94
91)	sec-Butylbenzene	17.17	105	191273	5.0866 ug/L		94
92)	p-Isopropyltoluene	17.32	119	159232	4.9597 ug/L		93
93)	1,3-Dichlorobenzene	17.52	146	83896	4.7972 ug/L		95
94)	1,4-Dichlorobenzene	17.64	146	82462	4.7242 ug/L		86
95)	n-Butylbenzene	17.85	91	158787	5.1996 ug/L		96
96)	1,2-Dichlorobenzene	18.14	146	72524	4.7321 ug/L		92
97)	1,2-Dibromo-3-Chloropropan	19.13	75	5629	4.6990 ug/L		76
98)	1,2,4-Trichlorobenzene	20.28	180	53855	4.5369 ug/L		98
99)	Hexachlorobutadiene	20.43	225	24504	4.4608 ug/L		93
100)	Naphthalene	20.65	128	79806	4.4217 ug/L		98
101)	1,2,3-Trichlorobenzene	20.96	180	44150	4.3720 ug/L		99

(#) = qualifier out of range (m) = manual integration
8M412224.D 8260WT.M Fri May 13 15:47:44 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412224.D Vial: 7
Acq On : 13 May 2016 15:25 Operator: TMB
Sample : WG568561-06 5ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 15:47 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Thu Apr 28 15:05:04 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412224.D Vial: 7
Acq On : 13 May 2016 15:25 Operator: TMB
Sample : WG568561-06 5ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	5.0000	4.3438	13.1	100	0.00
3	P	Chloromethane	5.0000	4.5471	9.1	100	0.00
4	C	Vinyl Chloride	5.0000	4.7641	4.7	100	0.00
5	T	1,3-Butadiene	5.0000	4.2777	14.4	100	0.00
6	T	Bromomethane	5.0000	4.1880	16.2	100	-0.01
7	T	Chloroethane	5.0000	4.4813	10.4	100	0.00
8	T	Trichlorofluoromethane	5.0000	4.6812	6.4	100	0.00
9	T	Diethyl ether	50.0000	47.5652	4.9	100	0.00
10	T	Isoprene	5.0000	4.4611	10.8	100	0.00
11	T	Acrolein	25.0000	21.7328	13.1	100	0.02
12	T	1,1,2-Trichloro-1,2,2-Trifl	5.0000	4.7244	5.5	100	0.00
13	T	Acetone	5.0000	4.1387	17.2	100	0.00
14	C	1,1-Dichloroethene	5.0000	4.7810	4.4	100	0.00
15	T	Tert-Butyl Alcohol	100.0000	100.4650	-0.5	100	0.00
16	T	Dimethyl Sulfide	5.0000	4.7499	5.0	100	0.00
17	T	Iodomethane	5.0000	3.7634	24.7	100	0.00
18	T	Methyl acetate	5.0000	4.9211	1.6	100	0.00
19	T	Methylene Chloride	5.0000	4.7467	5.1	100	0.00
20	T	Carbon Disulfide	5.0000	4.7328	5.3	100	0.00
21	T	Acrylonitrile	25.0000	22.8698	8.5	100	0.00
22	T	Methyl Tert Butyl Ether	5.0000	4.6750	6.5	100	0.00
23	T	trans-1,2-Dichloroethene	5.0000	4.7415	5.2	100	0.00
24	T	n-Hexane	5.0000	4.8084	3.8	100	0.00
25	T	Diisopropyl ether	50.0000	47.2872	5.4	100	0.00
26	T	Vinyl Acetate	5.0000	4.5766	8.5	100	0.00
27	P	1,1-Dichloroethane	5.0000	4.7058	5.9	100	0.00
28	T	Ethyl-Tert-Butyl ether	50.0000	47.5984	4.8	100	0.00
29	T	2-Butanone	5.0000	4.2191	15.6	100	0.00
30	T	Propionitrile	50.0000	48.4475	3.1	100	0.00
31	T	2,2-Dichloropropane	5.0000	4.8965	2.1	100	0.00
32	T	cis-1,2-Dichloroethene	5.0000	4.6525	7.0	100	0.00
33	C	Chloroform	5.0000	4.6735	6.5	100	0.00
34		1-Bromopropane	5.0000	4.8774	2.5	100	0.00
35	T	Bromochloromethane	5.0000	4.6403	7.2	100	0.00
36	T	Tetrahydrofuran	50.0000	45.5400	8.9	100	-0.01
37	S	Dibromofluoromethane	2.5000	2.3953	4.2	100	0.00
38	T	1,1,1-Trichloroethane	5.0000	4.7896	4.2	100	0.00
39	T	Cyclohexane	5.0000	4.6730	6.5	100	0.00
40	T	1,1-Dichloropropene	5.0000	4.6741	6.5	100	0.00
41	T	Tert-Amyl-Methyl ether	50.0000	47.6211	4.8	100	0.00
42	T	Carbon Tetrachloride	5.0000	4.8044	3.9	100	0.00
43	S	1,2-Dichloroethane-d4	2.5000	2.3949	4.2	100	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	5.0000	4.7996	4.0	100	0.00
46	T	Benzene	5.0000	4.7588	4.8	100	0.00
47	T	Trichloroethene	5.0000	4.5733	8.5	100	0.00
48	T	Methylcyclohexane	5.0000	4.7616	4.8	100	0.00
49	C	1,2-Dichloropropane	5.0000	4.5219	9.6	100	0.00
50	T	Bromodichloromethane	5.0000	4.7878	4.2	100	0.00
51	T	1,4-Dioxane	100.0000	77.0346	23.0	100	0.00
52	T	Dibromomethane	5.0000	4.7390	5.2	100	0.00
53	T	2-Chloroethyl Vinyl Ether	5.0000	4.2867	14.3	100	0.00
54	T	4-Methyl-2-Pentanone	5.0000	4.1576	16.8	100	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412224.D Vial: 7
Acq On : 13 May 2016 15:25 Operator: TMB
Sample : WG568561-06 5ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	5.0000	4.8662	2.7	100	0.00
56	T	Dimethyl Disulfide	5.0000	4.1683	16.6	100	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
58	S	Toluene-d8	2.5000	2.3473	6.1	100	0.00
59	C	Toluene	5.0000	4.8533	2.9	100	0.00
60	T	Ethyl Methacrylate	5.0000	4.4883	10.2	100	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	5.0000	4.4786	10.4	100	0.00
63	T	1,1,2-Trichloroethane	5.0000	4.5231	9.5	100	0.00
64	T	2-Hexanone	5.0000	4.0898	18.2	100	0.00
65	T	1,3-Dichloropropane	5.0000	4.8109	3.8	100	0.00
66	T	Tetrachloroethene	5.0000	4.6028	7.9	100	0.00
67	T	Dibromochloromethane	5.0000	4.5802	8.4	100	0.00
68	T	1,2-Dibromoethane	5.0000	4.6475	7.1	100	0.00
69	T	1-Chlorohexane	5.0000	4.7331	5.3	100	0.00
70	P	Chlorobenzene	5.0000	4.8107	3.8	100	0.00
71	T	1,1,1,2-Tetrachloroethane	5.0000	4.3984	12.0	100	0.00
72	C	Ethylbenzene	5.0000	4.6492	7.0	100	0.00
73	T	m-,p-Xylene	10.0000	9.5700	4.3	100	0.00
74	T	o-Xylene	5.0000	4.6773	6.5	100	0.00
75	T	Styrene	5.0000	4.8405	3.2	100	0.00
76	P	Bromoform	5.0000	4.2672	14.7	100	0.00
77	T	Isopropylbenzene	5.0000	4.8951	2.1	100	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
79	P	1,1,2,2-Tetrachloroethane	5.0000	4.8422	3.2	100	0.00
80	S	p-Bromofluorobenzene	2.5000	2.3565	5.7	100	0.00
81	T	1,2,3-Trichloropropane	5.0000	5.2150	-4.3	100	0.00
82	T	trans-1,4-Dichloro-2-Butene	5.0000	4.3781	12.4	100	0.00
83	T	n-Propylbenzene	5.0000	5.1935	-3.9	100	0.00
84	T	Bromobenzene	5.0000	4.8483	3.0	100	0.00
85	T	1,3,5-Trimethylbenzene	5.0000	4.9964	0.1	100	0.00
86	T	2-Chlorotoluene	5.0000	4.9803	0.4	100	0.00
87	T	4-Chlorotoluene	5.0000	5.0286	-0.6	100	0.00
88	T	a-Methylstyrene	5.0000	4.6629	6.7	100	0.00
89	T	tert-Butylbenzene	5.0000	4.7677	4.6	100	0.00
90	T	1,2,4-Trimethylbenzene	5.0000	5.0245	-0.5	100	0.00
91	T	sec-Butylbenzene	5.0000	5.2170	-4.3	100	0.00
92	T	p-Isopropyltoluene	5.0000	5.1150	-2.3	100	0.00
93	T	1,3-Dichlorobenzene	5.0000	4.9511	1.0	100	0.00
94	T	1,4-Dichlorobenzene	5.0000	4.8270	3.5	100	0.00
95	T	n-Butylbenzene	5.0000	5.1595	-3.2	100	0.00
96	T	1,2-Dichlorobenzene	5.0000	4.8122	3.8	100	0.00
97	T	1,2-Dibromo-3-Chloropropane	5.0000	5.0942	-1.9	100	0.00
98	T	1,2,4-Trichlorobenzene	5.0000	4.7808	4.4	100	0.00
99	T	Hexachlorobutadiene	5.0000	4.6555	6.9	100	0.00
100	T	Naphthalene	5.0000	4.7684	4.6	100	0.00
101	T	1,2,3-Trichlorobenzene	5.0000	4.6455	7.1	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412224.D 8260WT.M Sat May 14 18:08:37 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412225.D Vial: 8
Acq On : 13 May 2016 15:54 Operator: TMB
Sample : WG568561-07 20ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:16:32 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	770027	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	531490	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	283594	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	86295	10.9569 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 43.84%#		
43) 1,2-Dichloroethane-d4	10.30	65	96239	11.3450 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 45.36%#		
58) Toluene-d8	12.68	98	294491	10.9223 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 43.68%#		
80) p-Bromofluorobenzene	16.08	95	121459	10.5490 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 42.20%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	288235	23.7135 ug/L	99
3) Chloromethane	3.62	50	155712	20.3784 ug/L	98
4) Vinyl Chloride	3.84	62	184250	20.9368 ug/L	95
5) 1,3-Butadiene	3.89	54	149348	22.7750 ug/L	96
6) Bromomethane	4.73	94	100098	25.0143 ug/L	99
7) Chloroethane	4.89	64	100876	20.9404 ug/L	91
8) Trichlorofluoromethane	5.37	101	351299	21.2512 ug/L	98
9) Diethyl ether	5.90	59	324317	80.8265 ug/L	78
10) Isoprene	5.94	67	205842	20.3622 ug/L	77
11) Acrolein	6.14	56	12049	34.8201 ug/L	90
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	159996	20.3859 ug/L	97
13) Acetone	6.25	43	18459	25.7405 ug/L	98
14) 1,1-Dichloroethene	6.47	61	256241	20.5065 ug/L	92
15) Tert-Butyl Alcohol	6.57	59	43922	153.5712 ug/L	93
16) Dimethyl Sulfide	6.73	62	89456	20.6087 ug/L	83
17) Iodomethane	6.99	142	40622	61.0474 ug/L	93
18) Methyl acetate	7.00	43	46598	20.2171 ug/L #	85
19) Methylene Chloride	7.25	84	161897	20.1338 ug/L	52
20) Carbon Disulfide	7.30	76	538090	20.0802 ug/L	98
21) Acrylonitrile	7.43	53	46305	46.9889 ug/L	100
22) Methyl Tert Butyl Ether	7.47	73	361116	22.0137 ug/L	86
23) trans-1,2-Dichloroethene	7.70	61	234754	20.3175 ug/L	90
24) n-Hexane	7.78	57	172498	21.1165 ug/L	84
25) Diisopropyl ether	8.12	45	1447822	80.3755 ug/L #	91
26) Vinyl Acetate	8.30	43	166501	21.7017 ug/L #	88
27) 1,1-Dichloroethane	8.33	63	300321	21.3849 ug/L	97
28) Ethyl-Tert-Butyl ether	8.70	59	1579265	83.0733 ug/L	89
29) 2-Butanone	8.88	43	25681	23.9099 ug/L	87
30) Propionitrile	8.98	54	30953	89.5673 ug/L	94
31) 2,2-Dichloropropane	9.10	77	309575	20.8527 ug/L	100
32) cis-1,2-Dichloroethene	9.17	96	181973	20.4488 ug/L	89
33) Chloroform	9.37	83	348925	20.5535 ug/L	96
34) 1-Bromopropane	9.52	122	26723	22.7206 ug/L	96
35) Bromochloromethane	9.61	130	94952	21.3181 ug/L	65
36) Tetrahydrofuran	9.63	42	63066	70.5772 ug/L #	57
38) 1,1,1-Trichloroethane	9.91	97	335486	21.3640 ug/L	99
39) Cyclohexane	9.94	56	221040	19.9984 ug/L #	64
40) 1,1-Dichloropropene	10.11	75	248373	20.4717 ug/L	96
41) Tert-Amyl-Methyl ether	10.21	73	1418183	81.7883 ug/L	88
42) Carbon Tetrachloride	10.25	117	296471	21.3449 ug/L	98

(#) = qualifier out of range (m) = manual integration
8M412225.D 8260WT.M Fri May 13 16:16:35 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412225.D Vial: 8
Acq On : 13 May 2016 15:54 Operator: TMB
Sample : WG568561-07 20ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:16:32 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc	Unit		Qvalue
45)	1,2-Dichloroethane	10.43	62	228665	21.6403	ug/L		85
46)	Benzene	10.46	78	668449	20.3769	ug/L		86
47)	Trichloroethene	11.21	130	174107	20.7980	ug/L		96
48)	Methylcyclohexane	11.29	83	262275	21.4887	ug/L		82
49)	1,2-Dichloropropane	11.43	63	146868	21.1883	ug/L	#	72
50)	Bromodichloromethane	11.73	83	262192	21.9986	ug/L		98
51)	1,4-Dioxane	11.71	88	5351	196.3837	ug/L		76
52)	Dibromomethane	11.81	93	86951	21.0734	ug/L		98
53)	2-Chloroethyl Vinyl Ether	12.03	63	63586	22.4600	ug/L		91
54)	4-Methyl-2-Pentanone	12.05	58	25754	25.0531	ug/L	#	73
55)	cis-1,3-Dichloropropene	12.36	75	265439	21.9841	ug/L		99
56)	Dimethyl Disulfide	12.63	79	134192	23.9412	ug/L		100
59)	Toluene	12.78	91	700473	20.1722	ug/L		99
60)	Ethyl Methacrylate	12.88	69	133361	23.2046	ug/L		71
62)	trans-1,3-Dichloropropene	12.96	75	238585	23.2548	ug/L		99
63)	1,1,2-Trichloroethane	13.18	97	107901	21.0216	ug/L		99
64)	2-Hexanone	13.11	58	22561	25.2093	ug/L	#	85
65)	1,3-Dichloropropane	13.49	76	198804	22.2491	ug/L		74
66)	Tetrachloroethene	13.60	164	141791	20.0055	ug/L		94
67)	Dibromochloromethane	13.87	129	158465	22.0695	ug/L		99
68)	1,2-Dibromoethane	14.13	107	110085	22.3366	ug/L		100
69)	1-Chlorohexane	14.21	91	229753	20.3591	ug/L		83
70)	Chlorobenzene	14.62	112	467492	20.4613	ug/L		97
71)	1,1,1,2-Tetrachloroethane	14.65	131	183079	22.2095	ug/L		96
72)	Ethylbenzene	14.65	106	249016	20.2289	ug/L		90
73)	m-,p-Xylene	14.75	106	606381	41.1308	ug/L		88
74)	o-Xylene	15.30	106	294106	21.2106	ug/L		90
75)	Styrene	15.35	104	492781	21.8207	ug/L		89
76)	Bromoform	15.83	173	95911	23.8070	ug/L		99
77)	Isopropylbenzene	15.73	105	806678	20.7280	ug/L		97
79)	1,1,2,2-Tetrachloroethane	15.95	83	123253	22.7779	ug/L		95
81)	1,2,3-Trichloropropane	16.14	110	36712	22.1354	ug/L		94
82)	trans-1,4-Dichloro-2-Buten	16.18	53	33702	27.6513	ug/L	#	56
83)	n-Propylbenzene	16.24	91	962196	20.3511	ug/L		97
84)	Bromobenzene	16.37	156	190566	21.2033	ug/L		94
85)	1,3,5-Trimethylbenzene	16.43	105	692633	20.9615	ug/L		94
86)	2-Chlorotoluene	16.51	91	649077	19.9155	ug/L		96
87)	4-Chlorotoluene	16.56	91	595446	20.3225	ug/L		95
88)	a-Methylstyrene	16.83	118	318960	22.3680	ug/L		92
89)	tert-Butylbenzene	16.90	134	129440	20.4705	ug/L		83
90)	1,2,4-Trimethylbenzene	16.95	105	707749	20.4783	ug/L		94
91)	sec-Butylbenzene	17.17	105	818359	20.1802	ug/L		97
92)	p-Isopropyltoluene	17.32	119	677859	19.7492	ug/L		96
93)	1,3-Dichlorobenzene	17.52	146	364065	20.5126	ug/L		92
94)	1,4-Dichlorobenzene	17.64	146	363066	19.9874	ug/L		92
95)	n-Butylbenzene	17.85	91	681055	20.0187	ug/L		97
96)	1,2-Dichlorobenzene	18.14	146	318612	19.9358	ug/L		91
97)	1,2-Dibromo-3-Chloropropan	19.13	75	23070	21.6419	ug/L		93
98)	1,2,4-Trichlorobenzene	20.28	180	242455	20.8863	ug/L		98
99)	Hexachlorobutadiene	20.43	225	110384	20.2523	ug/L		96
100)	Naphthalene	20.65	128	369849	21.4526	ug/L		99
101)	1,2,3-Trichlorobenzene	20.96	180	198876	19.8356	ug/L		98

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412225.D Vial: 8
Acq On : 13 May 2016 15:54 Operator: TMB
Sample : WG568561-07 20ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:16 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412226.D Vial: 9
Acq On : 13 May 2016 16:23 Operator: TMB
Sample : WG568561-08 50ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:45:30 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	778260	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	539116	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	288225	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.67	111	211053	26.5139 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 106.04%		
43) 1,2-Dichloroethane-d4	10.31	65	229348	26.7504 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 107.00%		
58) Toluene-d8	12.68	98	718491	26.2709 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 105.08%		
80) p-Bromofluorobenzene	16.07	95	292755	25.0180 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 100.08%		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.17	85	682745	55.5761 ug/L	100
3) Chloromethane	3.62	50	372858	48.2806 ug/L	98
4) Vinyl Chloride	3.85	62	453905	51.0328 ug/L	96
5) 1,3-Butadiene	3.88	54	390501	58.9199 ug/L	99
6) Bromomethane	4.74	94	260493	64.4080 ug/L	97
7) Chloroethane	4.89	64	244336	50.1840 ug/L	91
8) Trichlorofluoromethane	5.38	101	837676	50.1376 ug/L	99
9) Diethyl ether	5.90	59	389327	96.0019 ug/L	79
10) Isoprene	5.94	67	498060	48.7477 ug/L	76
11) Acrolein	6.13	56	17644	50.4496 ug/L	99
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	387643	48.8691 ug/L	97
13) Acetone	6.25	43	45901	63.3305 ug/L	97
14) 1,1-Dichloroethene	6.46	61	619744	49.0724 ug/L	94
15) Tert-Butyl Alcohol	6.58	59	53656	185.6211 ug/L	91
16) Dimethyl Sulfide	6.73	62	221107	50.3993 ug/L	82
17) Iodomethane	6.98	142	157872	234.7428 ug/L	94
18) Methyl acetate	7.00	43	114759	49.2629 ug/L #	81
19) Methylene Chloride	7.25	84	389653	47.9453 ug/L	52
20) Carbon Disulfide	7.29	76	1269941	46.8897 ug/L	99
21) Acrylonitrile	7.43	53	59882	60.1236 ug/L	97
22) Methyl Tert Butyl Ether	7.46	73	901668	54.3843 ug/L	86
23) trans-1,2-Dichloroethene	7.70	61	571507	48.9396 ug/L	93
24) n-Hexane	7.79	57	419656	50.8290 ug/L	83
25) Diisopropyl ether	8.13	45	1781012	97.8265 ug/L #	90
26) Vinyl Acetate	8.29	43	428708	55.2866 ug/L #	89
27) 1,1-Dichloroethane	8.32	63	722574	50.9078 ug/L	96
28) Ethyl-Tert-Butyl ether	8.71	59	1955100	101.7552 ug/L	89
29) 2-Butanone	8.88	43	66579	61.3316 ug/L #	82
30) Propionitrile	8.98	54	38235	109.4684 ug/L	95
31) 2,2-Dichloropropane	9.10	77	768496	51.2175 ug/L	99
32) cis-1,2-Dichloroethene	9.17	96	439590	48.8754 ug/L	91
33) Chloroform	9.38	83	844564	49.2230 ug/L	97
34) 1-Bromopropane	9.51	122	63442	53.3693 ug/L	99
35) Bromochloromethane	9.60	130	230119	51.1185 ug/L	66
36) Tetrahydrofuran	9.64	42	79437	87.9575 ug/L #	62
38) 1,1,1-Trichloroethane	9.90	97	817606	51.5150 ug/L	99
39) Cyclohexane	9.93	56	535882	47.9707 ug/L #	65
40) 1,1-Dichloropropene	10.10	75	615439	50.1899 ug/L	96
41) Tert-Amyl-Methyl ether	10.21	73	1769381	100.9628 ug/L	88
42) Carbon Tetrachloride	10.24	117	724058	51.5782 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M412226.D 8260WT.M Fri May 13 16:45:32 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412226.D Vial: 9
Acq On : 13 May 2016 16:23 Operator: TMB
Sample : WG568561-08 50ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:45:30 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
45)	1,2-Dichloroethane	10.42	62	561107	52.5401 ug/L		85
46)	Benzene	10.46	78	1607407	48.4816 ug/L		86
47)	Trichloroethene	11.22	130	424998	50.2312 ug/L		96
48)	Methylcyclohexane	11.30	83	635339	51.5040 ug/L		83
49)	1,2-Dichloropropane	11.42	63	351908	50.2320 ug/L	#	71
50)	Bromodichloromethane	11.72	83	654943	54.3701 ug/L		97
51)	1,4-Dioxane	11.71	88	7573	274.9918 ug/L		73
52)	Dibromomethane	11.82	93	209167	50.1574 ug/L		97
53)	2-Chloroethyl Vinyl Ether	12.02	63	162319	56.7281 ug/L		93
54)	4-Methyl-2-Pentanone	12.05	58	64928	62.4928 ug/L		80
55)	cis-1,3-Dichloropropene	12.36	75	666861	54.6463 ug/L		98
56)	Dimethyl Disulfide	12.62	79	343446	60.6260 ug/L		96
59)	Toluene	12.79	91	1676069	47.5846 ug/L		100
60)	Ethyl Methacrylate	12.88	69	338996	58.1503 ug/L		69
62)	trans-1,3-Dichloropropene	12.96	75	594749	57.1499 ug/L		100
63)	1,1,2-Trichloroethane	13.18	97	268432	51.5570 ug/L		98
64)	2-Hexanone	13.12	58	56859	62.6347 ug/L	#	83
65)	1,3-Dichloropropane	13.48	76	489226	53.9770 ug/L		75
66)	Tetrachloroethene	13.60	164	350332	48.7296 ug/L		95
67)	Dibromochloromethane	13.86	129	401149	55.0780 ug/L		100
68)	1,2-Dibromoethane	14.12	107	273580	54.7252 ug/L		99
69)	1-Chlorohexane	14.21	91	560883	48.9984 ug/L		82
70)	Chlorobenzene	14.63	112	1117812	48.2326 ug/L		96
71)	1,1,1,2-Tetrachloroethane	14.66	131	446892	53.4462 ug/L		99
72)	Ethylbenzene	14.66	106	603074	48.2980 ug/L		91
73)	m-,p-Xylene	14.74	106	1443780	96.5463 ug/L		90
74)	o-Xylene	15.31	106	709244	50.4265 ug/L		90
75)	Styrene	15.34	104	1212099	52.9133 ug/L		91
76)	Bromoform	15.84	173	238354	58.3272 ug/L		100
77)	Isopropylbenzene	15.73	105	1919993	48.6372 ug/L		97
79)	1,1,2,2-Tetrachloroethane	15.95	83	294266	53.5083 ug/L		97
81)	1,2,3-Trichloropropane	16.14	110	89279	52.9656 ug/L		93
82)	trans-1,4-Dichloro-2-Buten	16.19	53	86829	70.0956 ug/L	#	53
83)	n-Propylbenzene	16.24	91	2278994	47.4277 ug/L		98
84)	Bromobenzene	16.37	156	459959	50.3549 ug/L		93
85)	1,3,5-Trimethylbenzene	16.43	105	1661486	49.4745 ug/L		95
86)	2-Chlorotoluene	16.51	91	1611031	48.6367 ug/L		97
87)	4-Chlorotoluene	16.56	91	1341051	45.0344 ug/L		95
88)	a-Methylstyrene	16.83	118	772611	53.3111 ug/L		90
89)	tert-Butylbenzene	16.89	134	307503	47.8493 ug/L		81
90)	1,2,4-Trimethylbenzene	16.94	105	1679809	47.8234 ug/L		95
91)	sec-Butylbenzene	17.16	105	1925411	46.7166 ug/L		97
92)	p-Isopropyltoluene	17.32	119	1621383	46.4795 ug/L		95
93)	1,3-Dichlorobenzene	17.51	146	882656	48.9327 ug/L		93
94)	1,4-Dichlorobenzene	17.64	146	871210	47.1909 ug/L		92
95)	n-Butylbenzene	17.85	91	1608832	46.5296 ug/L		98
96)	1,2-Dichlorobenzene	18.14	146	768548	47.3159 ug/L		91
97)	1,2-Dibromo-3-Chloropropan	19.13	75	59586	54.9992 ug/L		86
98)	1,2,4-Trichlorobenzene	20.28	180	583880	49.4902 ug/L		100
99)	Hexachlorobutadiene	20.44	225	264699	47.7844 ug/L		98
100)	Naphthalene	20.65	128	894629	51.0579 ug/L		99
101)	1,2,3-Trichlorobenzene	20.96	180	475246	46.6387 ug/L		99

(#) = qualifier out of range (m) = manual integration
8M412226.D 8260WT.M Fri May 13 16:45:33 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412226.D Vial: 9
Acq On : 13 May 2016 16:23 Operator: TMB
Sample : WG568561-08 50ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 16:45 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 15:56:03 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412227.D Vial: 10
Acq On : 13 May 2016 16:52 Operator: TMB
Sample : WG568561-09 100ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:14:31 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:12:43 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	809533	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	559075	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	296193	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	460283	53.9056 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 215.64%#		
43) 1,2-Dichloroethane-d4	10.30	65	488116	52.5821 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 210.32%#		
58) Toluene-d8	12.68	98	1540572	52.8075 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 211.24%#		
80) p-Bromofluorobenzene	16.08	95	623735	51.1573 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 204.64%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	1463283	109.1068 ug/L	100
3) Chloromethane	3.62	50	778702	97.2380 ug/L	98
4) Vinyl Chloride	3.84	62	968312	103.4979 ug/L	95
5) 1,3-Butadiene	3.88	54	896569	120.4972 ug/L	95
6) Bromomethane	4.73	94	605228	129.8677 ug/L	99
7) Chloroethane	4.89	64	528153	103.4126 ug/L	90
8) Trichlorofluoromethane	5.37	101	1782726	101.4754 ug/L	99
9) Diethyl ether	5.90	59	828091	197.4770 ug/L	79
10) Isoprene	5.94	67	1080684	101.8037 ug/L	78
11) Acrolein	6.14	56	38655	109.5572 ug/L	98
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	833546	101.0906 ug/L	97
13) Acetone	6.24	43	97515	109.1944 ug/L	100
14) 1,1-Dichloroethene	6.46	61	1323999	100.6729 ug/L	93
15) Tert-Butyl Alcohol	6.57	59	114220	387.5328 ug/L	91
16) Dimethyl Sulfide	6.73	62	489571	106.5996 ug/L	82
17) Iodomethane	6.99	142	416890	244.5557 ug/L	97
18) Methyl acetate	7.00	43	252444	104.3161 ug/L #	81
19) Methylene Chloride	7.24	84	827241	98.5346 ug/L	53
20) Carbon Disulfide	7.29	76	2766110	99.3432 ug/L	100
21) Acrylonitrile	7.43	53	126153	113.2272 ug/L	98
22) Methyl Tert Butyl Ether	7.46	73	1894494	105.8643 ug/L	87
23) trans-1,2-Dichloroethene	7.70	61	1203701	99.1822 ug/L	93
24) n-Hexane	7.78	57	926226	105.3097 ug/L	84
25) Diisopropyl ether	8.12	45	3699807	196.0386 ug/L #	90
26) Vinyl Acetate	8.29	43	860820	101.8643 ug/L #	88
27) 1,1-Dichloroethane	8.32	63	1521237	101.5567 ug/L	96
28) Ethyl-Tert-Butyl ether	8.70	59	4005676	198.2271 ug/L	89
29) 2-Butanone	8.88	43	145126	112.6698 ug/L #	81
30) Propionitrile	8.98	54	83038	219.1648 ug/L	97
31) 2,2-Dichloropropane	9.11	77	1610615	102.0557 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	935490	99.9944 ug/L	89
33) Chloroform	9.37	83	1766988	98.8341 ug/L	98
34) 1-Bromopropane	9.51	122	139657	108.5299 ug/L	100
35) Bromochloromethane	9.61	130	488678	102.8480 ug/L	66
36) Tetrahydrofuran	9.63	42	167812	187.5699 ug/L #	61
38) 1,1,1-Trichloroethane	9.91	97	1724504	102.7713 ug/L	100
39) Cyclohexane	9.94	56	1185537	102.8628 ug/L #	66
40) 1,1-Dichloropropene	10.11	75	1293224	100.8377 ug/L	96
41) Tert-Amyl-Methyl ether	10.21	73	3634806	198.1264 ug/L	88
42) Carbon Tetrachloride	10.25	117	1534696	103.3979 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M412227.D 8260WT.M Fri May 13 17:14:33 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412227.D Vial: 10
Acq On : 13 May 2016 16:52 Operator: TMB
Sample : WG568561-09 100ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:14:31 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:12:43 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc	Unit		Qvalue
45)	1,2-Dichloroethane	10.42	62	1177876	103.7352	ug/L		85
46)	Benzene	10.46	78	3298928	95.8405	ug/L		87
47)	Trichloroethene	11.21	130	903490	101.9036	ug/L		97
48)	Methylcyclohexane	11.30	83	1389528	104.6454	ug/L		83
49)	1,2-Dichloropropane	11.43	63	755128	102.5297	ug/L		73
50)	Bromodichloromethane	11.73	83	1351612	104.6036	ug/L		99
51)	1,4-Dioxane	11.72	88	16303	473.9624	ug/L		68
52)	Dibromomethane	11.81	93	445130	101.6543	ug/L		99
53)	2-Chloroethyl Vinyl Ether	12.02	63	333977	105.4228	ug/L		92
54)	4-Methyl-2-Pentanone	12.05	58	138980	110.1494	ug/L	#	77
55)	cis-1,3-Dichloropropene	12.36	75	1387475	105.9136	ug/L		100
56)	Dimethyl Disulfide	12.62	79	745504	111.3169	ug/L		97
59)	Toluene	12.78	91	3455614	95.2346	ug/L		99
60)	Ethyl Methacrylate	12.88	69	731009	113.5760	ug/L		70
62)	trans-1,3-Dichloropropene	12.96	75	1236182	107.9445	ug/L		99
63)	1,1,2-Trichloroethane	13.18	97	564076	103.0604	ug/L		99
64)	2-Hexanone	13.12	58	126268	114.5366	ug/L	#	79
65)	1,3-Dichloropropane	13.48	76	1021082	105.2672	ug/L		74
66)	Tetrachloroethene	13.60	164	752835	101.4023	ug/L		95
67)	Dibromochloromethane	13.87	129	856499	109.6525	ug/L		100
68)	1,2-Dibromoethane	14.12	107	577701	107.6426	ug/L		98
69)	1-Chlorohexane	14.21	91	1222720	103.0385	ug/L		82
70)	Chlorobenzene	14.62	112	2332247	97.2408	ug/L		96
71)	1,1,1,2-Tetrachloroethane	14.66	131	947925	106.1463	ug/L		99
72)	Ethylbenzene	14.66	106	1275519	98.8771	ug/L		97
73)	m-,p-Xylene	14.75	106	3009424	194.2602	ug/L		98
74)	o-Xylene	15.31	106	1490638	100.8068	ug/L		93
75)	Styrene	15.35	104	2503721	102.8373	ug/L		88
76)	Bromoform	15.83	173	517827	114.0520	ug/L		100
77)	Isopropylbenzene	15.73	105	3913141	95.4433	ug/L		99
79)	1,1,2,2-Tetrachloroethane	15.95	83	624965	106.8610	ug/L		97
81)	1,2,3-Trichloropropane	16.14	110	192087	107.3268	ug/L		97
82)	trans-1,4-Dichloro-2-Buten	16.18	53	189214	128.4818	ug/L	#	59
83)	n-Propylbenzene	16.24	91	4599057	93.6644	ug/L		100
84)	Bromobenzene	16.37	156	968011	102.1426	ug/L		95
85)	1,3,5-Trimethylbenzene	16.43	105	3401719	97.9556	ug/L		97
86)	2-Chlorotoluene	16.52	91	3177047	93.8267	ug/L		98
87)	4-Chlorotoluene	16.56	91	2899171	96.0712	ug/L		96
88)	a-Methylstyrene	16.84	118	1639166	103.6811	ug/L		90
89)	tert-Butylbenzene	16.90	134	654025	99.4199	ug/L		85
90)	1,2,4-Trimethylbenzene	16.95	105	3479295	96.7054	ug/L		97
91)	sec-Butylbenzene	17.17	105	3947832	94.2785	ug/L		98
92)	p-Isopropyltoluene	17.32	119	3328450	94.4148	ug/L		97
93)	1,3-Dichlorobenzene	17.52	146	1834740	98.9074	ug/L		92
94)	1,4-Dichlorobenzene	17.64	146	1818309	96.6271	ug/L		92
95)	n-Butylbenzene	17.85	91	3308330	94.4002	ug/L		98
96)	1,2-Dichlorobenzene	18.14	146	1606254	97.0179	ug/L		91
97)	1,2-Dibromo-3-Chloropropan	19.13	75	125368	107.7020	ug/L		87
98)	1,2,4-Trichlorobenzene	20.28	180	1227926	100.7076	ug/L		99
99)	Hexachlorobutadiene	20.43	225	569213	100.5231	ug/L		97
100)	Naphthalene	20.65	128	1852486	101.2968	ug/L		99
101)	1,2,3-Trichlorobenzene	20.96	180	1021163	98.5794	ug/L		100

(#) = qualifier out of range (m) = manual integration
8M412227.D 8260WT.M Fri May 13 17:14:34 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412227.D Vial: 10
Acq On : 13 May 2016 16:52 Operator: TMB
Sample : WG568561-09 100ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:14 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:12:43 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412228.D Vial: 11
Acq On : 13 May 2016 17:21 Operator: TMB
Sample : WG568561-10 200ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:43:32 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	828540	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	578848	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	301401	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	898145	101.4517 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 405.80%#		
43) 1,2-Dichloroethane-d4	10.30	65	939848	98.0779 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 392.32%#		
58) Toluene-d8	12.68	98	2984162	97.8805 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 391.52%#		
80) p-Bromofluorobenzene	16.08	95	1242720	99.7023 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 398.80%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	2830517	203.5621 ug/L	98
3) Chloromethane	3.62	50	1586058	194.4056 ug/L	97
4) Vinyl Chloride	3.83	62	1935009	201.0738 ug/L	94
5) 1,3-Butadiene	3.88	54	1684964	212.5477 ug/L	94
6) Bromomethane	4.72	94	1358068	271.2231 ug/L	99
7) Chloroethane	4.89	64	1084803	206.5255 ug/L	90
8) Trichlorofluoromethane	5.37	101	3516760	195.1756 ug/L	99
10) Isoprene	5.94	67	2136665	196.1573 ug/L	78
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	1662142	196.5993 ug/L	95
13) Acetone	6.24	43	187748	200.7963 ug/L	97
14) 1,1-Dichloroethene	6.46	61	2649821	196.6732 ug/L	92
15) Tert-Butyl Alcohol	6.58	59	6024	20.0950 ug/L #	87
16) Dimethyl Sulfide	6.73	62	976862	205.8822 ug/L	82
17) Iodomethane	6.99	142	896947	398.7987 ug/L	100
18) Methyl acetate	7.00	43	504293	201.4324 ug/L #	81
19) Methylene Chloride	7.25	84	1684780	196.5546 ug/L	52
20) Carbon Disulfide	7.29	76	5386210	189.2120 ug/L	99
21) Acrylonitrile	7.45	53	30354	26.0447 ug/L #	32
22) Methyl Tert Butyl Ether	7.46	73	3711011	200.6529 ug/L	87
23) trans-1,2-Dichloroethene	7.70	61	2410498	194.2902 ug/L	92
24) n-Hexane	7.78	57	1821777	199.7288 ug/L	84
26) Vinyl Acetate	8.29	43	1796217	206.9060 ug/L #	88
27) 1,1-Dichloroethane	8.32	63	3042055	197.9863 ug/L	95
29) 2-Butanone	8.88	43	271767	199.8193 ug/L #	77
30) Propionitrile	8.90	54	208	0.5280 ug/L #	1
31) 2,2-Dichloropropane	9.11	77	3202633	197.6970 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	1875631	195.8883 ug/L	89
33) Chloroform	9.37	83	3506648	191.9199 ug/L	99
34) 1-Bromopropane	9.51	122	286158	214.2313 ug/L	99
35) Bromochloromethane	9.60	130	980373	200.7807 ug/L	66
36) Tetrahydrofuran	9.62	42	2381	2.6275 ug/L	87
38) 1,1,1-Trichloroethane	9.91	97	3410306	197.7909 ug/L	99
39) Cyclohexane	9.94	56	2357393	198.8974 ug/L #	66
40) 1,1-Dichloropropene	10.10	75	2601175	197.8946 ug/L	96
42) Carbon Tetrachloride	10.25	117	3062767	200.6417 ug/L	99
45) 1,2-Dichloroethane	10.42	62	2313662	198.0324 ug/L	85
46) Benzene	10.46	78	6379817	182.1771 ug/L	89
47) Trichloroethene	11.21	130	1810459	198.9741 ug/L	97
48) Methylcyclohexane	11.30	83	2755373	200.4193 ug/L	83
49) 1,2-Dichloropropane	11.43	63	1528407	202.0331 ug/L	74

(#) = qualifier out of range (m) = manual integration
8M412228.D 8260WT.M Fri May 13 17:43:34 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412228.D Vial: 11
Acq On : 13 May 2016 17:21 Operator: TMB
Sample : WG568561-10 200ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:43:32 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
50)	Bromodichloromethane	11.73	83	2716764	204.0895 ug/L		99
51)	1,4-Dioxane	11.73	88	3865	104.9352 ug/L	#	18
52)	Dibromomethane	11.81	93	898971	200.1154 ug/L		98
53)	2-Chloroethyl Vinyl Ether	12.02	63	631656	192.7236 ug/L		90
54)	4-Methyl-2-Pentanone	12.05	58	262702	198.3957 ug/L	#	75
55)	cis-1,3-Dichloropropene	12.36	75	2784162	205.9153 ug/L		99
56)	Dimethyl Disulfide	12.62	79	1518705	215.4711 ug/L		95
59)	Toluene	12.78	91	6564353	175.9274 ug/L		96
60)	Ethyl Methacrylate	12.88	69	1429837	209.8161 ug/L		71
62)	trans-1,3-Dichloropropene	12.96	75	2468609	205.4771 ug/L		98
63)	1,1,2-Trichloroethane	13.18	97	1124672	197.6019 ug/L		100
64)	2-Hexanone	13.12	58	237678	200.9290 ug/L	#	81
65)	1,3-Dichloropropane	13.48	76	2023247	199.9545 ug/L		76
66)	Tetrachloroethene	13.60	164	1546618	200.8018 ug/L		97
67)	Dibromochloromethane	13.87	129	1726523	210.5823 ug/L		100
68)	1,2-Dibromoethane	14.12	107	1163160	207.0667 ug/L		99
69)	1-Chlorohexane	14.21	91	2414147	195.6413 ug/L		83
70)	Chlorobenzene	14.62	112	4641169	187.6386 ug/L		94
71)	1,1,1,2-Tetrachloroethane	14.66	131	1929251	206.8372 ug/L		99
72)	Ethylbenzene	14.66	106	2611321	195.8267 ug/L		88
73)	m-,p-Xylene	14.75	106	5876234	367.8657 ug/L		83
74)	o-Xylene	15.31	106	3012318	196.4899 ug/L		98
75)	Styrene	15.35	104	4913204	194.1235 ug/L		86
76)	Bromoform	15.83	173	1052629	218.7989 ug/L		100
77)	Isopropylbenzene	15.73	105	7284624	172.7305 ug/L		97
79)	1,1,2,2-Tetrachloroethane	15.95	83	1237171	205.8673 ug/L		98
81)	1,2,3-Trichloropropane	16.14	110	386738	209.7906 ug/L	#	54
82)	trans-1,4-Dichloro-2-Buten	16.19	53	382467	243.6526 ug/L		76
83)	n-Propylbenzene	16.24	91	8244387	166.5110 ug/L		93
84)	Bromobenzene	16.37	156	1935166	200.1305 ug/L		95
85)	1,3,5-Trimethylbenzene	16.43	105	6482975	183.9948 ug/L		98
86)	2-Chlorotoluene	16.52	91	6108308	178.8549 ug/L		99
87)	4-Chlorotoluene	16.56	91	5513221	180.5508 ug/L		99
88)	a-Methylstyrene	16.84	118	3255254	200.4995 ug/L		89
89)	tert-Butylbenzene	16.90	134	1351569	202.1005 ug/L		92
90)	1,2,4-Trimethylbenzene	16.95	105	6531416	179.2445 ug/L		97
91)	sec-Butylbenzene	17.17	105	7342730	173.9814 ug/L		97
92)	p-Isopropyltoluene	17.32	119	6352010	178.7315 ug/L		99
93)	1,3-Dichlorobenzene	17.52	146	3640547	193.1654 ug/L		90
94)	1,4-Dichlorobenzene	17.64	146	3593548	188.4603 ug/L		91
95)	n-Butylbenzene	17.85	91	6198060	175.4375 ug/L		97
96)	1,2-Dichlorobenzene	18.14	146	3171699	188.9651 ug/L		89
97)	1,2-Dibromo-3-Chloropropan	19.13	75	249646	207.5645 ug/L		88
98)	1,2,4-Trichlorobenzene	20.28	180	2480626	199.7296 ug/L		99
99)	Hexachlorobutadiene	20.43	225	1169411	202.7981 ug/L		98
100)	Naphthalene	20.65	128	3580198	192.0323 ug/L		97
101)	1,2,3-Trichlorobenzene	20.96	180	2040570	193.9298 ug/L		100

(#) = qualifier out of range (m) = manual integration
8M412228.D 8260WT.M Fri May 13 17:43:35 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412228.D Vial: 11
Acq On : 13 May 2016 17:21 Operator: TMB
Sample : WG568561-10 200ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 17:43 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412229.D Vial: 12
Acq On : 13 May 2016 17:50 Operator: TMB
Sample : WG568561-11 300ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 18:12:33 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	855409	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	599314	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	305726	25.00 ug/L	0.00

System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	1411837	154.4674 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 617.88%#		
43) 1,2-Dichloroethane-d4	10.30	65	1466019	148.1811 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 592.72%#		
58) Toluene-d8	12.69	98	4594713	145.5601 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 582.24%#		
80) p-Bromofluorobenzene	16.08	95	1918996	151.7813 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 607.12%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	4191932	292.0014 ug/L	97
3) Chloromethane	3.62	50	2474035	293.7212 ug/L	97
4) Vinyl Chloride	3.83	62	2956347	297.5552 ug/L	94
5) 1,3-Butadiene	3.87	54	2605506	318.3447 ug/L	94
6) Bromomethane	4.72	94	2177550	421.2239 ug/L	98
7) Chloroethane	4.89	64	1688005	311.2693 ug/L	90
8) Trichlorofluoromethane	5.37	101	5387662	289.6161 ug/L	99
9) Diethyl ether	5.90	59	1306432	295.4610 ug/L	80
10) Isoprene	5.93	67	3377164	300.3032 ug/L	78
11) Acrolein	6.14	56	70572	185.7401 ug/L	97
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	2563878	293.7317 ug/L	94
13) Acetone	6.24	43	283185	293.3529 ug/L	90
14) 1,1-Dichloroethene	6.46	61	4031745	289.8421 ug/L	91
15) Tert-Butyl Alcohol	6.58	59	181796	587.3907 ug/L	90
16) Dimethyl Sulfide	6.73	62	1553493	317.1280 ug/L	81
17) Iodomethane	6.99	142	1347307	580.2206 ug/L	98
18) Methyl acetate	7.00	43	800046	309.5287 ug/L #	80
19) Methylene Chloride	7.24	84	2587731	292.4144 ug/L	52
20) Carbon Disulfide	7.29	76	8228181	279.9683 ug/L	98
21) Acrylonitrile	7.43	53	228227	189.6753 ug/L	94
22) Methyl Tert Butyl Ether	7.46	73	5605924	293.5892 ug/L	88
23) trans-1,2-Dichloroethene	7.70	61	3678699	287.1957 ug/L	92
24) n-Hexane	7.78	57	2863434	304.0691 ug/L	86
25) Diisopropyl ether	8.12	45	5819647	292.7898 ug/L #	89
26) Vinyl Acetate	8.30	43	2819503	314.5767 ug/L #	87
27) 1,1-Dichloroethane	8.32	63	4586531	289.1293 ug/L	94
28) Ethyl-Tert-Butyl ether	8.70	59	6190206	290.3321 ug/L	89
29) 2-Butanone	8.88	43	439359	312.8959 ug/L #	76
30) Propionitrile	8.99	54	133197	327.4673 ug/L	95
31) 2,2-Dichloropropane	9.10	77	4821388	288.2735 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	2847005	287.9976 ug/L	89
33) Chloroform	9.37	83	5197860	275.5447 ug/L	100
34) 1-Bromopropane	9.51	122	450012	326.3179 ug/L	100
35) Bromochloromethane	9.61	130	1490221	295.6112 ug/L	66
36) Tetrahydrofuran	9.63	42	271736	290.4491 ug/L #	62
38) 1,1,1-Trichloroethane	9.91	97	5126354	287.9792 ug/L	98
39) Cyclohexane	9.94	56	3733392	305.0987 ug/L #	67
40) 1,1-Dichloropropene	10.11	75	3923871	289.1471 ug/L	97
41) Tert-Amyl-Methyl ether	10.21	73	5639573	291.3713 ug/L	88
42) Carbon Tetrachloride	10.25	117	4598150	291.7629 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M412229.D 8260WT.M Fri May 13 18:12:35 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412229.D Vial: 12
Acq On : 13 May 2016 17:50 Operator: TMB
Sample : WG568561-11 300ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 18:12:33 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit	Qvalue
45)	1,2-Dichloroethane	10.43	62	3468738	287.5726 ug/L	86
46)	Benzene	10.46	78	9065861	250.7462 ug/L	93
47)	Trichloroethene	11.21	130	2772282	295.1106 ug/L	97
48)	Methylcyclohexane	11.30	83	4327357	304.8748 ug/L	84
49)	1,2-Dichloropropane	11.43	63	2350729	300.9718 ug/L	74
50)	Bromodichloromethane	11.73	83	4051493	294.7973 ug/L	99
51)	1,4-Dioxane	11.72	88	26614	699.8767 ug/L	67
52)	Dibromomethane	11.81	93	1369945	295.3776 ug/L	97
53)	2-Chloroethyl Vinyl Ether	12.03	63	1006317	297.3917 ug/L	90
54)	4-Methyl-2-Pentanone	12.06	58	424684	310.6522 ug/L #	76
55)	cis-1,3-Dichloropropene	12.36	75	4156310	297.7431 ug/L	97
56)	Dimethyl Disulfide	12.63	79	2359825	324.2911 ug/L	94
59)	Toluene	12.78	91	9135213	236.4669 ug/L	90
60)	Ethyl Methacrylate	12.88	69	2228037	315.7801 ug/L	71
62)	trans-1,3-Dichloropropene	12.96	75	3700845	297.5240 ug/L	96
63)	1,1,2-Trichloroethane	13.18	97	1730199	293.6104 ug/L	100
64)	2-Hexanone	13.11	58	381967	311.8816 ug/L #	78
65)	1,3-Dichloropropane	13.49	76	3080177	294.0141 ug/L	76
66)	Tetrachloroethene	13.60	164	2374435	297.7522 ug/L	96
67)	Dibromochloromethane	13.87	129	2638145	310.7838 ug/L	100
68)	1,2-Dibromoethane	14.13	107	1778561	305.8085 ug/L	99
69)	1-Chlorohexane	14.21	91	3712592	290.5924 ug/L	84
70)	Chlorobenzene	14.63	112	6754240	263.7433 ug/L	90
71)	1,1,1,2-Tetrachloroethane	14.65	131	2967650	307.3001 ug/L	100
72)	Ethylbenzene	14.65	106	4014412	290.7661 ug/L	70
73)	m-,p-Xylene	14.75	106	8357822	505.3514 ug/L	70
74)	o-Xylene	15.31	106	4530840	285.4488 ug/L	88
75)	Styrene	15.35	104	7050702	269.0641 ug/L	82
76)	Bromoform	15.84	173	1602587	321.7373 ug/L	99
77)	Isopropylbenzene	15.73	105	9900817	226.7477 ug/L	90
79)	1,1,2,2-Tetrachloroethane	15.95	83	1885953	309.3861 ug/L	98
81)	1,2,3-Trichloropropane	16.14	110	581198	310.8176 ug/L	61
82)	trans-1,4-Dichloro-2-Buten	16.18	53	589705	370.3601 ug/L	82
83)	n-Propylbenzene	16.25	91	10768378	214.4110 ug/L	86
84)	Bromobenzene	16.37	156	2947671	300.5290 ug/L	99
85)	1,3,5-Trimethylbenzene	16.43	105	8890651	248.7580 ug/L	92
86)	2-Chlorotoluene	16.51	91	7868844	227.1449 ug/L	92
87)	4-Chlorotoluene	16.57	91	7782731	251.2687 ug/L	93
88)	a-Methylstyrene	16.84	118	4955768	300.9205 ug/L	86
89)	tert-Butylbenzene	16.90	134	2046401	301.6701 ug/L	95
90)	1,2,4-Trimethylbenzene	16.95	105	8903769	240.8932 ug/L	89
91)	sec-Butylbenzene	17.17	105	9867862	230.5051 ug/L	91
92)	p-Isopropyltoluene	17.32	119	8738927	242.4155 ug/L	94
93)	1,3-Dichlorobenzene	17.52	146	5319954	278.2806 ug/L	87
94)	1,4-Dichlorobenzene	17.64	146	5248511	271.3593 ug/L	88
95)	n-Butylbenzene	17.85	91	8408157	234.6279 ug/L	92
96)	1,2-Dichlorobenzene	18.14	146	4717385	277.0787 ug/L	87
97)	1,2-Dibromo-3-Chloropropan	19.13	75	377070	309.0741 ug/L	91
98)	1,2,4-Trichlorobenzene	20.28	180	3677657	291.9206 ug/L	98
99)	Hexachlorobutadiene	20.43	225	1780695	304.4378 ug/L	96
100)	Naphthalene	20.65	128	5216751	275.8542 ug/L	95
101)	1,2,3-Trichlorobenzene	20.96	180	3064908	287.1593 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M412229.D 8260WT.M Fri May 13 18:12:36 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDchem\2\data\051316\8M412229.D Vial: 12
Acq On : 13 May 2016 17:50 Operator: TMB
Sample : WG568561-11 300ug/L STD 8260 Inst : HPMS8
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 13 18:12 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Fri May 13 17:17:39 2016
Response via : Initial Calibration



8M412229.D 8260WT.M Fri May 13 18:12:36 2016 Page 3



Method Name: K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:08:06 2016



Method Name: K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:08:06 2016



Method Name: K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:08:06 2016

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\2\DATA\051316\8M412232.D Vial: 15
Acq On : 13 May 2016 19:17 Operator: TMB
Sample : WG568561-12 50ug/L ALT SRC STD 8260 Inst : HPMS8
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: May 17 11:44:36 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Tue May 17 11:43:20 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	837879	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.58	117	574720	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	304324	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.67	111	229742	25.5204 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 102.08%		
43) 1,2-Dichloroethane-d4	10.31	65	238672	24.7259 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 98.92%		
58) Toluene-d8	12.68	98	776362	25.8115 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 103.24%		
80) p-Bromofluorobenzene	16.07	95	315683	25.0519 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 100.20%		
Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.17	85	761591	54.2141 ug/L	99
3) Chloromethane	3.62	50	415589	50.6814 ug/L	100
4) Vinyl Chloride	3.84	62	509607	52.3810 ug/L	100
5) 1,3-Butadiene	3.88	54	347562	42.6002 ug/L	97
6) Bromomethane	4.73	94	272060	46.3924 ug/L	99
7) Chloroethane	4.88	64	285509	53.3336 ug/L	98
8) Trichlorofluoromethane	5.37	101	849580	46.9313 ug/L	100
9) Diethyl ether	5.90	59	451307	104.4282 ug/L	99
10) Isoprene	5.93	67	582863	53.0208 ug/L	99
11) Acrolein	6.13	56	38877	100.4722 ug/L	99
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	419584	49.3091 ug/L	97
13) Acetone	6.23	43	54234	57.5310 ug/L	94
14) 1,1-Dichloroethene	6.46	61	622831	45.9700 ug/L	99
15) Tert-Butyl Alcohol	6.59	59	59797	197.9422 ug/L	97
16) Dimethyl Sulfide	6.73	62	351814	72.6234 ug/L	98
17) Iodomethane	6.99	142	92716	26.2832 ug/L	99
18) Methyl acetate	7.00	43	127540	50.0514 ug/L	98
19) Methylene Chloride	7.25	84	417642	48.4385 ug/L	99
20) Carbon Disulfide	7.29	76	1199899	42.3200 ug/L	100
21) Acrylonitrile	7.43	53	65324	53.4074 ug/L	93
22) Methyl Tert Butyl Ether	7.46	73	949572	50.8859 ug/L	97
23) trans-1,2-Dichloroethene	7.70	61	598296	48.0666 ug/L	99
24) n-Hexane	7.77	57	425106	45.9930 ug/L	99
25) Diisopropyl ether	8.13	45	2040382	105.1615 ug/L	98
26) Vinyl Acetate	8.29	43	415568	46.7802 ug/L	99
27) 1,1-Dichloroethane	8.32	63	721661	46.6846 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	2040020	98.1343 ug/L	99
29) 2-Butanone	8.88	43	73394	52.9905 ug/L	99
30) Propionitrile	8.98	54	42290	104.7756 ug/L	100
31) 2,2-Dichloropropane	9.10	77	780073	47.8861 ug/L	99
32) cis-1,2-Dichloroethene	9.16	96	470465	48.9162 ug/L	98
33) Chloroform	9.38	83	860252	47.1317 ug/L	100
34) 1-Bromopropane	9.51	122	98585	71.5615 ug/L	98
35) Bromochloromethane	9.60	130	234493	47.5456 ug/L	99
36) Tetrahydrofuran	9.62	42	86049	94.3281 ug/L	99
38) 1,1,1-Trichloroethane	9.90	97	825167	47.5949 ug/L	100
39) Cyclohexane	9.93	56	586721	48.8808 ug/L	97
40) 1,1-Dichloropropene	10.10	75	627217	47.4632 ug/L	99
41) Tert-Amyl-Methyl ether	10.21	73	1962875	103.9618 ug/L	99
42) Carbon Tetrachloride	10.24	117	711823	46.2363 ug/L	100

(#) = qualifier out of range (m) = manual integration
8M412232.D 8260WT.M Tue May 17 11:44:38 2016 Page 1

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412232.D Vial: 15
Acq On : 13 May 2016 19:17 Operator: TMB
Sample : WG568561-12 50ug/L ALT SRC STD 8260 Inst : HPMS8
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	108	0.00
2	T	Dichlorodifluoromethane	50.0000	54.2141	-8.4	112	0.00
3	P	Chloromethane	50.0000	50.6814	-1.4	111	0.00
4	C	Vinyl Chloride	50.0000	52.3811	-4.8	112	-0.01
5	T	1,3-Butadiene	50.0000	42.6002	14.8	89	0.00
6	T	Bromomethane	50.0000	46.3924	7.2	104	-0.01
7	T	Chloroethane	50.0000	53.3336	-6.7	117	-0.01
8	T	Trichlorofluoromethane	50.0000	46.9313	6.1	101	-0.01
9	T	Diethyl ether	100.0000	104.4282	-4.4	116	0.00
10	T	Isoprene	50.0000	53.0208	-6.0	117	0.00
11	T	Acrolein	50.0000	100.4722	-100.9#	220	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	50.0000	49.3091	1.4	108	0.00
13	T	Acetone	50.0000	57.5310	-15.1	118	-0.01
14	C	1,1-Dichloroethene	50.0000	45.9700	8.1	100	0.00
15	T	Tert-Butyl Alcohol	200.0000	197.9423	1.0	108	0.01
16	T	Dimethyl Sulfide	50.0000	72.6234	-45.2#	159	0.00
17	T	Iodomethane	50.0000	26.2832	47.4#	59	0.01
18	T	Methyl acetate	50.0000	50.0514	-0.1	111	0.00
19	T	Methylene Chloride	50.0000	48.4385	3.1	107	0.00
20	T	Carbon Disulfide	50.0000	42.3200	15.4	94	0.00
21	T	Acrylonitrile	50.0000	53.4074	-6.8	109	0.00
22	T	Methyl Tert Butyl Ether	50.0000	50.8859	-1.8	105	0.00
23	T	trans-1,2-Dichloroethene	50.0000	48.0666	3.9	105	0.00
24	T	n-Hexane	50.0000	45.9930	8.0	101	-0.01
25	T	Diisopropyl ether	100.0000	105.1615	-5.2	115	0.00
26	T	Vinyl Acetate	50.0000	46.7802	6.4	97	0.00
27	P	1,1-Dichloroethane	50.0000	46.6846	6.6	100	0.00
28	T	Ethyl-Tert-Butyl ether	100.0000	98.1343	1.9	104	0.00
29	T	2-Butanone	50.0000	52.9905	-6.0	110	0.00
30	T	Propionitrile	100.0000	104.7756	-4.8	111	0.00
31	T	2,2-Dichloropropane	50.0000	47.8861	4.2	102	0.00
32	T	cis-1,2-Dichloroethene	50.0000	48.9162	2.2	107	-0.01
33	C	Chloroform	50.0000	47.1317	5.7	102	0.00
34		1-Bromopropane	50.0000	71.5615	-43.1#	155	0.00
35	T	Bromochloromethane	50.0000	47.5456	4.9	102	0.00
36	T	Tetrahydrofuran	100.0000	94.3281	5.7	108	-0.01
37	S	Dibromofluoromethane	25.0000	25.5204	-2.1	109	0.00
38	T	1,1,1-Trichloroethane	50.0000	47.5949	4.8	101	0.00
39	T	Cyclohexane	50.0000	48.8808	2.2	109	0.00
40	T	1,1-Dichloropropene	50.0000	47.4632	5.1	102	0.00
41	T	Tert-Amyl-Methyl ether	100.0000	103.9618	-4.0	111	0.00
42	T	Carbon Tetrachloride	50.0000	46.2363	7.5	98	0.00
43	S	1,2-Dichloroethane-d4	25.0000	24.7259	1.1	104	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	50.0000	49.0148	2.0	103	0.00
46	T	Benzene	50.0000	48.5232	3.0	104	0.00
47	T	Trichloroethene	50.0000	51.0658	-2.1	110	0.00
48	T	Methylcyclohexane	50.0000	49.8573	0.3	109	0.00
49	C	1,2-Dichloropropane	50.0000	50.9747	-1.9	111	0.00
50	T	Bromodichloromethane	50.0000	48.3053	3.4	99	0.00
51	T	1,4-Dioxane	200.0000	216.9538	-8.5	110	0.01
52	T	Dibromomethane	50.0000	45.8126	8.4	99	0.00
53	T	2-Chloroethyl Vinyl Ether	50.0000	51.2623	-2.5	104	0.00
54	T	4-Methyl-2-Pentanone	50.0000	51.8715	-3.7	107	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\2\DATA\051316\8M412232.D Vial: 15
Acq On : 13 May 2016 19:17 Operator: TMB
Sample : WG568561-12 50ug/L ALT SRC STD 8260 Inst : HPMS8
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	50.0000	53.8979	-7.8	111	0.00
56	T	Dimethyl Disulfide	50.0000	52.2714	-4.5	111	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	107	0.00
58	S	Toluene-d8	25.0000	25.8115	-3.2	108	0.00
59	C	Toluene	50.0000	49.0912	1.8	105	0.00
60	T	Ethyl Methacrylate	50.0000	55.9509	-11.9	113	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	50.0000	50.6740	-1.3	102	0.00
63	T	1,1,2-Trichloroethane	50.0000	49.0377	1.9	103	0.00
64	T	2-Hexanone	50.0000	53.6057	-7.2	112	0.00
65	T	1,3-Dichloropropane	50.0000	53.6720	-7.3	110	0.00
66	T	Tetrachloroethene	50.0000	47.8053	4.4	104	0.00
67	T	Dibromochloromethane	50.0000	49.6585	0.7	102	0.00
68	T	1,2-Dibromoethane	50.0000	51.1121	-2.2	105	0.00
69	T	1-Chlorohexane	50.0000	51.0611	-2.1	111	0.00
70	P	Chlorobenzene	50.0000	50.1844	-0.4	108	0.00
71	T	1,1,1,2-Tetrachloroethane	50.0000	49.3793	1.2	103	0.00
72	C	Ethylbenzene	50.0000	48.2013	3.6	105	0.00
73	T	m-,p-Xylene	100.0000	101.1341	-1.1	108	0.00
74	T	o-Xylene	50.0000	50.5675	-1.1	108	0.00
75	T	Styrene	50.0000	51.5086	-3.0	105	0.00
76	P	Bromoform	50.0000	48.2227	3.6	99	0.00
77	T	Isopropylbenzene	50.0000	51.1140	-2.2	107	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	106	0.00
79	P	1,1,2,2-Tetrachloroethane	50.0000	49.6172	0.8	103	0.00
80	S	p-Bromofluorobenzene	25.0000	25.0519	-0.2	108	0.00
81	T	1,2,3-Trichloropropane	50.0000	50.0849	-0.2	106	0.01
82	T	trans-1,4-Dichloro-2-Butene	50.0000	39.2043	21.6	86	0.00
83	T	n-Propylbenzene	50.0000	51.9237	-3.8	108	0.00
84	T	Bromobenzene	50.0000	49.4648	1.1	105	0.00
85	T	1,3,5-Trimethylbenzene	50.0000	51.6344	-3.3	107	0.00
86	T	2-Chlorotoluene	50.0000	51.5639	-3.1	106	0.00
87	T	4-Chlorotoluene	50.0000	48.2650	3.5	108	0.00
88	T	a-Methylstyrene	50.0000	56.4132	-12.8	120	0.00
89	T	tert-Butylbenzene	50.0000	50.9494	-1.9	112	0.00
90	T	1,2,4-Trimethylbenzene	50.0000	49.6617	0.7	105	0.00
91	T	sec-Butylbenzene	50.0000	51.3386	-2.7	108	0.00
92	T	p-Isopropyltoluene	50.0000	50.1261	-0.3	107	0.00
93	T	1,3-Dichlorobenzene	50.0000	50.8943	-1.8	108	0.00
94	T	1,4-Dichlorobenzene	50.0000	50.4742	-0.9	110	0.00
95	T	n-Butylbenzene	50.0000	49.6520	0.7	105	0.00
96	T	1,2-Dichlorobenzene	50.0000	50.4907	-1.0	110	0.00
97	T	1,2-Dibromo-3-Chloropropane	50.0000	48.4851	3.0	100	0.00
98	T	1,2,4-Trichlorobenzene	50.0000	50.1129	-0.2	107	0.00
99	T	Hexachlorobutadiene	50.0000	52.0919	-4.2	115	0.00
100	T	Naphthalene	50.0000	45.8047	8.4	95	0.00
101	T	1,2,3-Trichlorobenzene	50.0000	47.7353	4.5	106	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M412232.D 8260WT.M Sat May 14 18:09:03 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413126.D Vial: 3
Acq On : 21 Jun 2016 18:23 Operator: FJB
Sample : WG573644-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 18:45:45 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	842468	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	634272	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	348189	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	242647	26.8071 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 107.24%		
43) 1,2-Dichloroethane-d4	10.30	65	262534	27.0498 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 108.20%		
58) Toluene-d8	12.68	98	848655	25.5659 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 102.28%		
80) p-Bromofluorobenzene	16.08	95	365799	25.3719 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 101.48%		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	706967	50.0516 ug/L	100
3) Chloromethane	3.62	50	386080	46.8263 ug/L	100
4) Vinyl Chloride	3.84	62	531769	54.3613 ug/L	100
5) 1,3-Butadiene	3.89	54	469338	57.2127 ug/L	96
6) Bromomethane	4.73	94	237465	40.5323 ug/L	99
7) Chloroethane	4.89	64	282748	52.5301 ug/L	98
8) Trichlorofluoromethane	5.37	101	1010155	55.4976 ug/L	100
9) Diethyl ether	5.90	59	465053	107.0227 ug/L	98
10) Isoprene	5.94	67	560373	50.6973 ug/L	99
11) Acrolein	6.14	56	17931	46.0877 ug/L	100
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	460994	53.8805 ug/L	94
13) Acetone	6.24	43	54713	57.7230 ug/L	99
14) 1,1-Dichloroethene	6.47	61	729079	53.5188 ug/L	100
15) Tert-Butyl Alcohol	6.57	59	48518	159.7312 ug/L	97
16) Dimethyl Sulfide	6.73	62	239591	49.1883 ug/L	100
17) Iodomethane	6.99	142	129268	35.2171 ug/L	98
18) Methyl acetate	7.00	43	133943	52.2779 ug/L	99
19) Methylene Chloride	7.24	84	441231	50.8956 ug/L	99
20) Carbon Disulfide	7.29	76	1520883	53.3488 ug/L	100
21) Acrylonitrile	7.43	53	66221	53.8458 ug/L	97
22) Methyl Tert Butyl Ether	7.46	73	1012281	53.9509 ug/L	100
23) trans-1,2-Dichloroethene	7.70	61	658832	52.6417 ug/L	99
24) n-Hexane	7.78	57	525323	56.5261 ug/L	100
25) Diisopropyl ether	8.12	45	2046398	104.8971 ug/L	99
26) Vinyl Acetate	8.30	43	419641	46.9814 ug/L	99
27) 1,1-Dichloroethane	8.32	63	823421	52.9774 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	2274881	108.8361 ug/L	98
29) 2-Butanone	8.88	43	67631	48.5636 ug/L	93
30) Propionitrile	8.98	54	41084	101.2332 ug/L	94
31) 2,2-Dichloropropane	9.10	77	966004	58.9767 ug/L	99
32) cis-1,2-Dichloroethene	9.17	96	505239	52.2457 ug/L	100
33) Chloroform	9.37	83	977628	53.2708 ug/L	100
34) 1-Bromopropane	9.51	122	79927	57.7019 ug/L	99
35) Bromochloromethane	9.61	130	268391	54.1223 ug/L	95
36) Tetrahydrofuran	9.63	42	81203	88.5309 ug/L	93
38) 1,1,1-Trichloroethane	9.91	97	970703	55.6843 ug/L	100
39) Cyclohexane	9.94	56	642456	53.2326 ug/L	98
40) 1,1-Dichloropropene	10.11	75	712738	53.6410 ug/L	99
41) Tert-Amyl-Methyl ether	10.21	73	2069642	109.0195 ug/L	99
42) Carbon Tetrachloride	10.25	117	866447	55.9733 ug/L	100

(#) = qualifier out of range (m) = manual integration
8M413126.D 8260WT.M Fri Jun 24 17:14:31 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413126.D Vial: 3
Acq On : 21 Jun 2016 18:23 Operator: FJB
Sample : WG573644-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 18:45:45 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
45)	1,2-Dichloroethane	10.42	62	648513	54.9030 ug/L		100
46)	Benzene	10.46	78	1791975	51.7816 ug/L		100
47)	Trichloroethene	11.21	130	497052	53.8524 ug/L		98
48)	Methylcyclohexane	11.30	83	772305	55.0785 ug/L		98
49)	1,2-Dichloropropane	11.43	63	387114	50.2500 ug/L		96
50)	Bromodichloromethane	11.73	83	732132	54.0715 ug/L		99
51)	1,4-Dioxane	11.72	88	5489	141.8410 ug/L		84
52)	Dibromomethane	11.81	93	231585	50.7834 ug/L		97
53)	2-Chloroethyl Vinyl Ether	12.03	63	149076	45.0223 ug/L		96
54)	4-Methyl-2-Pentanone	12.06	58	67090	49.6023 ug/L		95
55)	cis-1,3-Dichloropropene	12.36	75	738187	53.5621 ug/L		99
56)	Dimethyl Disulfide	12.62	79	379506	51.5921 ug/L		100
59)	Toluene	12.78	91	1909258	48.4871 ug/L		100
60)	Ethyl Methacrylate	12.88	69	368867	48.7782 ug/L		98
62)	trans-1,3-Dichloropropene	12.96	75	669685	50.7497 ug/L		99
63)	1,1,2-Trichloroethane	13.18	97	294334	47.3701 ug/L		99
64)	2-Hexanone	13.12	58	60649	46.4489 ug/L	#	98
65)	1,3-Dichloropropane	13.48	76	533207	48.1994 ug/L		100
66)	Tetrachloroethene	13.60	164	438543	51.9821 ug/L		94
67)	Dibromochloromethane	13.87	129	454999	50.1513 ug/L		99
68)	1,2-Dibromoethane	14.13	107	294945	47.6288 ug/L		100
69)	1-Chlorohexane	14.21	91	667148	49.6341 ug/L		99
70)	Chlorobenzene	14.62	112	1281159	48.2494 ug/L		99
71)	1,1,1,2-Tetrachloroethane	14.65	131	512855	49.8550 ug/L		99
72)	Ethylbenzene	14.65	106	690518	47.5308 ug/L		100
73)	m-,p-Xylene	14.75	106	1654045	97.0666 ug/L		100
74)	o-Xylene	15.31	106	807880	48.4927 ug/L		99
75)	Styrene	15.35	104	1355516	49.6076 ug/L		97
76)	Bromoform	15.83	173	282531	52.5027 ug/L		99
77)	Isopropylbenzene	15.73	105	2218306	50.1226 ug/L		99
79)	1,1,2,2-Tetrachloroethane	15.95	83	314446	44.9902 ug/L		99
81)	1,2,3-Trichloropropane	16.14	110	100648	46.7642 ug/L		89
82)	trans-1,4-Dichloro-2-Buten	16.18	53	98274	44.8606 ug/L		88
83)	n-Propylbenzene	16.24	91	2624924	48.3222 ug/L		99
84)	Bromobenzene	16.37	156	542381	48.5427 ug/L		95
85)	1,3,5-Trimethylbenzene	16.42	105	1926734	48.6921 ug/L		100
86)	2-Chlorotoluene	16.50	91	1832398	48.3152 ug/L		99
87)	4-Chlorotoluene	16.56	91	1589494	46.3979 ug/L		100
88)	a-Methylstyrene	16.83	118	856154	45.6045 ug/L		100
89)	tert-Butylbenzene	16.89	134	367638	47.4906 ug/L		98
90)	1,2,4-Trimethylbenzene	16.94	105	1954303	48.0312 ug/L		99
91)	sec-Butylbenzene	17.16	105	2214729	47.5764 ug/L		98
92)	p-Isopropyltoluene	17.32	119	1909463	48.3096 ug/L		99
93)	1,3-Dichlorobenzene	17.51	146	1043963	48.5233 ug/L		98
94)	1,4-Dichlorobenzene	17.64	146	1026577	47.3283 ug/L		99
95)	n-Butylbenzene	17.85	91	1846719	47.2606 ug/L		98
96)	1,2-Dichlorobenzene	18.14	146	894170	46.7296 ug/L		98
97)	1,2-Dibromo-3-Chloropropan	19.13	75	63774	45.4568 ug/L		97
98)	1,2,4-Trichlorobenzene	20.28	180	702261	49.0994 ug/L		100
99)	Hexachlorobutadiene	20.43	225	325396	48.6913 ug/L		98
100)	Naphthalene	20.65	128	959183	45.1381 ug/L		99
101)	1,2,3-Trichlorobenzene	20.96	180	560837	46.4781 ug/L		98

(#) = qualifier out of range (m) = manual integration
8M413126.D 8260WT.M Fri Jun 24 17:14:32 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413126.D Vial: 3
Acq On : 21 Jun 2016 18:23 Operator: FJB
Sample : WG573644-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 18:45 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413126.D 8260WT.M Fri Jun 24 17:14:33 2016 Page 3

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413126.D Vial: 3
Acq On : 21 Jun 2016 18:23 Operator: FJB
Sample : WG573644-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	1.0000	1.0000	0.0	108	0.00
2	T	Dichlorodifluoromethane	0.4192	0.4196	-0.1	104	0.00
3	P	Chloromethane	0.2447	0.2291	6.3	104	0.00
4	C	Vinyl Chloride	0.2903	0.3156	-8.7	117	0.00
5	T	1,3-Butadiene	0.2434	0.2786	-14.4	120	0.00
6	T	Bromomethane	0.1654	0.1409	14.8	91	0.00
7	T	Chloroethane	0.1597	0.1678	-5.1	116	0.00
8	T	Trichlorofluoromethane	0.5401	0.5995	-11.0	121	0.00
9	T	Diethyl ether	0.1290	0.1380	-7.0	119	0.00
10	T	Isoprene	0.3280	0.3326	-1.4	113	0.00
11	T	Acrolein	0.0116	0.0106	7.9	102	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	0.2539	0.2736	-7.8	119	0.00
13	T	Acetone	0.0281	0.0325	-15.4	119	0.00
14	C	1,1-Dichloroethene	0.4043	0.4327	-7.0	118	0.00
15	T	Tert-Butyl Alcohol	0.0090	0.0072	20.1	87	0.00
16	T	Dimethyl Sulfide	0.1445	0.1422	1.6	108	0.00
17	T	Iodomethane	0.0866	0.0767	11.4	82	0.00
18	T	Methyl acetate	0.0760	0.0795	-4.6	117	0.00
19	T	Methylene Chloride	0.2573	0.2619	-1.8	113	0.00
20	T	Carbon Disulfide	0.8460	0.9026	-6.7	120	0.00
21	T	Acrylonitrile	0.0365	0.0393	-7.7	111	0.00
22	T	Methyl Tert Butyl Ether	0.5568	0.6008	-7.9	112	0.00
23	T	trans-1,2-Dichloroethene	0.3714	0.3910	-5.3	115	0.00
24	T	n-Hexane	0.2758	0.3118	-13.1	125	0.00
25	T	Diisopropyl ether	0.5789	0.6073	-4.9	115	0.00
26	T	Vinyl Acetate	0.2651	0.2490	6.0	98	0.00
27	P	1,1-Dichloroethane	0.4612	0.4887	-6.0	114	0.00
28	T	Ethyl-Tert-Butyl ether	0.6203	0.6751	-8.8	116	0.00
29	T	2-Butanone	0.0413	0.0401	2.9	102	0.00
30	T	Propionitrile	0.0120	0.0122	-1.2	107	0.00
31	T	2,2-Dichloropropane	0.4860	0.5733	-18.0	126	0.00
32	T	cis-1,2-Dichloroethene	0.2870	0.2999	-4.5	115	0.00
33	C	Chloroform	0.5446	0.5802	-6.5	116	0.00
34		1-Bromopropane	0.0411	0.0474	-15.4	126	0.00
35	T	Bromochloromethane	0.1472	0.1593	-8.2	117	0.00
36	T	Tetrahydrofuran	0.0272	0.0241	11.5	102	0.00
37	S	Dibromofluoromethane	0.2686	0.2880	-7.2	115	0.00
38	T	1,1,1-Trichloroethane	0.5173	0.5761	-11.4	119	0.00
39	T	Cyclohexane	0.3581	0.3813	-6.5	120	0.00
40	T	1,1-Dichloropropene	0.3943	0.4230	-7.3	116	0.00
41	T	Tert-Amyl-Methyl ether	0.5634	0.6142	-9.0	117	0.00
42	T	Carbon Tetrachloride	0.4593	0.5142	-11.9	120	0.00
43	S	1,2-Dichloroethane-d4	0.2880	0.3116	-8.2	114	0.00
44		Heptane	0.0000	0.0000	0.0	0#	-2.54#
45	T	1,2-Dichloroethane	0.3505	0.3849	-9.8	116	0.00
46	T	Benzene	1.0269	1.0635	-3.6	111	0.00
47	T	Trichloroethene	0.2739	0.2950	-7.7	117	0.00
48	T	Methylcyclohexane	0.4161	0.4584	-10.2	122	0.00
49	C	1,2-Dichloropropane	0.2286	0.2298	-0.5	110	0.00
50	T	Bromodichloromethane	0.4018	0.4345	-8.1	112	0.00
51	T	1,4-Dioxane	0.0012	0.0008	29.6#	72	0.00
52	T	Dibromomethane	0.1353	0.1374	-1.6	111	0.00
53	T	2-Chloroethyl Vinyl Ether	0.0983	0.0885	10.0	92	0.00
54	T	4-Methyl-2-Pentanone	0.0401	0.0398	0.8	103	0.00

(#) = Out of Range

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413126.D Vial: 3
Acq On : 21 Jun 2016 18:23 Operator: FJB
Sample : WG573644-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	0.4090	0.4381	-7.1	111	0.00
56	T	Dimethyl Disulfide	0.2183	0.2252	-3.2	110	0.00
57	I	Chlorobenzene-d5	1.0000	1.0000	0.0	118	0.00
58	S	Toluene-d8	1.3084	1.3380	-2.3	118	0.00
59	C	Toluene	1.5520	1.5051	3.0	114	0.00
60	T	Ethyl Methacrylate	0.2981	0.2908	2.4	109	0.00
61		Paraldehyde	0.0000	0.0000	0.0	0#	-13.14#
62	T	trans-1,3-Dichloropropene	0.5201	0.5279	-1.5	113	0.00
63	T	1,1,2-Trichloroethane	0.2449	0.2320	5.3	110	0.00
64	T	2-Hexanone	0.0515	0.0478	7.1	107	0.00
65	T	1,3-Dichloropropane	0.4360	0.4203	3.6	109	0.00
66	T	Tetrachloroethene	0.3325	0.3457	-4.0	125	0.00
67	T	Dibromochloromethane	0.3576	0.3587	-0.3	113	0.00
68	T	1,2-Dibromoethane	0.2441	0.2325	4.7	108	0.00
69	T	1-Chlorohexane	0.5298	0.5259	0.7	119	0.00
70	P	Chlorobenzene	1.0466	1.0099	3.5	115	0.00
71	T	1,1,1,2-Tetrachloroethane	0.4055	0.4043	0.3	115	0.00
72	C	Ethylbenzene	0.5726	0.5443	4.9	114	0.00
73	T	m-,p-Xylene	0.6716	0.6520	2.9	115	0.00
74	T	o-Xylene	0.6566	0.6369	3.0	114	0.00
75	T	Styrene	1.0770	1.0686	0.8	112	0.00
76	P	Bromoform	0.2121	0.2227	-5.0	119	0.00
77	T	Isopropylbenzene	1.7444	1.7487	-0.2	116	0.00
78	I	1,4-Dichlorobenzene-d4	1.0000	1.0000	0.0	121	0.00
79	P	1,1,2,2-Tetrachloroethane	0.5018	0.4515	10.0	107	0.00
80	S	p-Bromofluorobenzene	1.0352	1.0506	-1.5	125	0.00
81	T	1,2,3-Trichloropropane	0.1545	0.1445	6.5	113	0.00
82	T	trans-1,4-Dichloro-2-Butene	0.1376	0.1411	-2.6	113	0.00
83	T	n-Propylbenzene	3.9003	3.7694	3.4	115	0.00
84	T	Bromobenzene	0.8022	0.7789	2.9	118	0.00
85	T	1,3,5-Trimethylbenzene	2.8411	2.7668	2.6	116	0.00
86	T	2-Chlorotoluene	2.7231	2.6313	3.4	114	0.00
87	T	4-Chlorotoluene	2.4597	2.2825	7.2	119	0.00
88	T	a-Methylstyrene	1.3479	1.2294	8.8	111	0.00
89	T	tert-Butylbenzene	0.5558	0.5279	5.0	120	0.00
90	T	1,2,4-Trimethylbenzene	2.9214	2.8064	3.9	116	0.00
91	T	sec-Butylbenzene	3.3424	3.1803	4.8	115	0.00
92	T	p-Isopropyltoluene	2.8379	2.7420	3.4	118	0.00
93	T	1,3-Dichlorobenzene	1.5448	1.4991	3.0	118	0.00
94	T	1,4-Dichlorobenzene	1.5574	1.4742	5.3	118	0.00
95	T	n-Butylbenzene	2.8056	2.6519	5.5	115	0.00
96	T	1,2-Dichlorobenzene	1.3739	1.2840	6.5	116	0.00
97	T	1,2-Dibromo-3-Chloropropane	0.1007	0.0916	9.1	107	0.00
98	T	1,2,4-Trichlorobenzene	1.0269	1.0085	1.8	120	0.00
99	T	Hexachlorobutadiene	0.4798	0.4673	2.6	123	0.00
100	T	Naphthalene	1.5257	1.3774	9.7	107	0.00
101	T	1,2,3-Trichlorobenzene	0.8664	0.8054	7.0	118	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413126.D Vial: 3
Acq On : 21 Jun 2016 18:23 Operator: FJB
Sample : WG573644-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	108	0.00
2	T	Dichlorodifluoromethane	50.0000	50.0516	-0.1	104	0.00
3	P	Chloromethane	50.0000	46.8263	6.3	104	0.00
4	C	Vinyl Chloride	50.0000	54.3613	-8.7	117	0.00
5	T	1,3-Butadiene	50.0000	57.2127	-14.4	120	0.00
6	T	Bromomethane	50.0000	40.5323	18.9	91	0.00
7	T	Chloroethane	50.0000	52.5301	-5.1	116	0.00
8	T	Trichlorofluoromethane	50.0000	55.4976	-11.0	121	0.00
9	T	Diethyl ether	100.0000	107.0227	-7.0	119	0.00
10	T	Isoprene	50.0000	50.6973	-1.4	113	0.00
11	T	Acrolein	50.0000	46.0877	7.8	102	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	50.0000	53.8805	-7.8	119	0.00
13	T	Acetone	50.0000	57.7230	-15.4	119	0.00
14	C	1,1-Dichloroethene	50.0000	53.5189	-7.0	118	0.00
15	T	Tert-Butyl Alcohol	200.0000	159.7312	20.1	87	0.00
16	T	Dimethyl Sulfide	50.0000	49.1883	1.6	108	0.00
17	T	Iodomethane	50.0000	35.2171	29.6#	82	0.00
18	T	Methyl acetate	50.0000	52.2779	-4.6	117	0.00
19	T	Methylene Chloride	50.0000	50.8956	-1.8	113	0.00
20	T	Carbon Disulfide	50.0000	53.3488	-6.7	120	0.00
21	T	Acrylonitrile	50.0000	53.8458	-7.7	111	0.00
22	T	Methyl Tert Butyl Ether	50.0000	53.9509	-7.9	112	0.00
23	T	trans-1,2-Dichloroethene	50.0000	52.6417	-5.3	115	0.00
24	T	n-Hexane	50.0000	56.5261	-13.1	125	0.00
25	T	Diisopropyl ether	100.0000	104.8971	-4.9	115	0.00
26	T	Vinyl Acetate	50.0000	46.9814	6.0	98	0.00
27	P	1,1-Dichloroethane	50.0000	52.9774	-6.0	114	0.00
28	T	Ethyl-Tert-Butyl ether	100.0000	108.8361	-8.8	116	0.00
29	T	2-Butanone	50.0000	48.5636	2.9	102	0.00
30	T	Propionitrile	100.0000	101.2332	-1.2	107	0.00
31	T	2,2-Dichloropropane	50.0000	58.9767	-18.0	126	0.00
32	T	cis-1,2-Dichloroethene	50.0000	52.2457	-4.5	115	0.00
33	C	Chloroform	50.0000	53.2708	-6.5	116	0.00
34		1-Bromopropane	50.0000	57.7019	-15.4	126	0.00
35	T	Bromochloromethane	50.0000	54.1223	-8.2	117	0.00
36	T	Tetrahydrofuran	100.0000	88.5309	11.5	102	0.00
37	S	Dibromofluoromethane	25.0000	26.8071	-7.2	115	0.00
38	T	1,1,1-Trichloroethane	50.0000	55.6843	-11.4	119	0.00
39	T	Cyclohexane	50.0000	53.2326	-6.5	120	0.00
40	T	1,1-Dichloropropene	50.0000	53.6410	-7.3	116	0.00
41	T	Tert-Amyl-Methyl ether	100.0000	109.0195	-9.0	117	0.00
42	T	Carbon Tetrachloride	50.0000	55.9733	-11.9	120	0.00
43	S	1,2-Dichloroethane-d4	25.0000	27.0498	-8.2	114	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	50.0000	54.9030	-9.8	116	0.00
46	T	Benzene	50.0000	51.7816	-3.6	111	0.00
47	T	Trichloroethene	50.0000	53.8524	-7.7	117	0.00
48	T	Methylcyclohexane	50.0000	55.0785	-10.2	122	0.00
49	C	1,2-Dichloropropane	50.0000	50.2500	-0.5	110	0.00
50	T	Bromodichloromethane	50.0000	54.0715	-8.1	112	0.00
51	T	1,4-Dioxane	200.0000	141.8410	29.1#	72	0.00
52	T	Dibromomethane	50.0000	50.7834	-1.6	111	0.00
53	T	2-Chloroethyl Vinyl Ether	50.0000	45.0223	10.0	92	0.00
54	T	4-Methyl-2-Pentanone	50.0000	49.6023	0.8	103	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413126.D Vial: 3
Acq On : 21 Jun 2016 18:23 Operator: FJB
Sample : WG573644-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	50.0000	53.5621	-7.1	111	0.00
56	T	Dimethyl Disulfide	50.0000	51.5921	-3.2	110	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	118	0.00
58	S	Toluene-d8	25.0000	25.5659	-2.3	118	0.00
59	C	Toluene	50.0000	48.4871	3.0	114	0.00
60	T	Ethyl Methacrylate	50.0000	48.7782	2.4	109	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	50.0000	50.7497	-1.5	113	0.00
63	T	1,1,2-Trichloroethane	50.0000	47.3701	5.3	110	0.00
64	T	2-Hexanone	50.0000	46.4489	7.1	107	0.00
65	T	1,3-Dichloropropane	50.0000	48.1994	3.6	109	0.00
66	T	Tetrachloroethene	50.0000	51.9821	-4.0	125	0.00
67	T	Dibromochloromethane	50.0000	50.1513	-0.3	113	0.00
68	T	1,2-Dibromoethane	50.0000	47.6288	4.7	108	0.00
69	T	1-Chlorohexane	50.0000	49.6341	0.7	119	0.00
70	P	Chlorobenzene	50.0000	48.2494	3.5	115	0.00
71	T	1,1,1,2-Tetrachloroethane	50.0000	49.8550	0.3	115	0.00
72	C	Ethylbenzene	50.0000	47.5308	4.9	114	0.00
73	T	m-,p-Xylene	100.0000	97.0666	2.9	115	0.00
74	T	o-Xylene	50.0000	48.4927	3.0	114	0.00
75	T	Styrene	50.0000	49.6076	0.8	112	0.00
76	P	Bromoform	50.0000	52.5027	-5.0	119	0.00
77	T	Isopropylbenzene	50.0000	50.1226	-0.2	116	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	121	0.00
79	P	1,1,2,2-Tetrachloroethane	50.0000	44.9902	10.0	107	0.00
80	S	p-Bromofluorobenzene	25.0000	25.3719	-1.5	125	0.00
81	T	1,2,3-Trichloropropane	50.0000	46.7642	6.5	113	0.00
82	T	trans-1,4-Dichloro-2-Butene	50.0000	44.8606	10.3	113	0.00
83	T	n-Propylbenzene	50.0000	48.3222	3.4	115	0.00
84	T	Bromobenzene	50.0000	48.5427	2.9	118	0.00
85	T	1,3,5-Trimethylbenzene	50.0000	48.6921	2.6	116	0.00
86	T	2-Chlorotoluene	50.0000	48.3152	3.4	114	0.00
87	T	4-Chlorotoluene	50.0000	46.3979	7.2	119	0.00
88	T	a-Methylstyrene	50.0000	45.6045	8.8	111	0.00
89	T	tert-Butylbenzene	50.0000	47.4906	5.0	120	0.00
90	T	1,2,4-Trimethylbenzene	50.0000	48.0312	3.9	116	0.00
91	T	sec-Butylbenzene	50.0000	47.5763	4.8	115	0.00
92	T	p-Isopropyltoluene	50.0000	48.3096	3.4	118	0.00
93	T	1,3-Dichlorobenzene	50.0000	48.5234	3.0	118	0.00
94	T	1,4-Dichlorobenzene	50.0000	47.3283	5.3	118	0.00
95	T	n-Butylbenzene	50.0000	47.2606	5.5	115	0.00
96	T	1,2-Dichlorobenzene	50.0000	46.7296	6.5	116	0.00
97	T	1,2-Dibromo-3-Chloropropane	50.0000	45.4568	9.1	107	0.00
98	T	1,2,4-Trichlorobenzene	50.0000	49.0994	1.8	120	0.00
99	T	Hexachlorobutadiene	50.0000	48.6914	2.6	123	0.00
100	T	Naphthalene	50.0000	45.1381	9.7	107	0.00
101	T	1,2,3-Trichlorobenzene	50.0000	46.4781	7.0	118	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M413126.D 8260WT.M Fri Jun 24 17:18:59 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413163.D Vial: 3
Acq On : 22 Jun 2016 15:41 Operator: FJB
Sample : WG573770-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 16:05:40 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	733044	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	548013	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	310940	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	213894	27.1579 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 108.64%		
43) 1,2-Dichloroethane-d4	10.30	65	238865	28.2849 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 113.12%		
58) Toluene-d8	12.68	98	756784	26.3867 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 105.56%		
80) p-Bromofluorobenzene	16.08	95	326001	25.3202 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 101.28%		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	662713	53.9222 ug/L	100
3) Chloromethane	3.62	50	356220	49.6540 ug/L	100
4) Vinyl Chloride	3.84	62	489912	57.5583 ug/L	99
5) 1,3-Butadiene	3.89	54	474280	66.4454 ug/L	94
6) Bromomethane	4.73	94	186080	36.6673 ug/L	100
7) Chloroethane	4.89	64	247391	52.8222 ug/L	98
8) Trichlorofluoromethane	5.37	101	959949	60.6119 ug/L	100
9) Diethyl ether	5.90	59	374507	99.0505 ug/L	98
10) Isoprene	5.94	67	563283	58.5676 ug/L	97
11) Acrolein	6.14	56	12754	37.6748 ug/L	87
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	409083	54.9504 ug/L	95
13) Acetone	6.23	43	48270	58.5273 ug/L	97
14) 1,1-Dichloroethene	6.46	61	674118	56.8711 ug/L	97
15) Tert-Butyl Alcohol	6.56	59	46245	174.9747 ug/L	99
16) Dimethyl Sulfide	6.73	62	213259	50.3179 ug/L	100
17) Iodomethane	6.99	142	144082	44.1276 ug/L	97
18) Methyl acetate	7.00	43	124376	55.7902 ug/L	98
19) Methylene Chloride	7.25	84	383069	50.7825 ug/L	100
20) Carbon Disulfide	7.29	76	1450754	58.4851 ug/L	99
21) Acrylonitrile	7.43	53	55427	51.7966 ug/L	99
22) Methyl Tert Butyl Ether	7.46	73	880555	53.9358 ug/L	100
23) trans-1,2-Dichloroethene	7.70	61	591515	54.3180 ug/L	100
24) n-Hexane	7.78	57	463490	57.3174 ug/L	99
25) Diisopropyl ether	8.12	45	1751539	103.1850 ug/L	96
26) Vinyl Acetate	8.30	43	337574	43.4350 ug/L	99
27) 1,1-Dichloroethane	8.32	63	746209	55.1763 ug/L	100
28) Ethyl-Tert-Butyl ether	8.70	59	1993570	109.6148 ug/L	99
29) 2-Butanone	8.88	43	58450	48.2362 ug/L	94
30) Propionitrile	8.99	54	33973	96.2072 ug/L	99
31) 2,2-Dichloropropane	9.10	77	907242	63.6573 ug/L	100
32) cis-1,2-Dichloroethene	9.17	96	448932	53.3528 ug/L	98
33) Chloroform	9.37	83	889950	55.7319 ug/L	100
34) 1-Bromopropane	9.51	122	76625	63.5756 ug/L	99
35) Bromochloromethane	9.60	130	235050	54.4744 ug/L	96
36) Tetrahydrofuran	9.63	42	72506	90.8490 ug/L	98
38) 1,1,1-Trichloroethane	9.91	97	905263	59.6822 ug/L	99
39) Cyclohexane	9.93	56	580560	55.2847 ug/L	97
40) 1,1-Dichloropropene	10.10	75	652272	56.4182 ug/L	100
41) Tert-Amyl-Methyl ether	10.21	73	1799715	108.9522 ug/L	100
42) Carbon Tetrachloride	10.25	117	820447	60.9134 ug/L	99

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413163.D Vial: 3
Acq On : 22 Jun 2016 15:41 Operator: FJB
Sample : WG573770-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 16:05:40 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit	Qvalue
45)	1,2-Dichloroethane	10.42	62	596728	58.0600 ug/L	98
46)	Benzene	10.46	78	1604181	53.2746 ug/L	99
47)	Trichloroethene	11.21	130	444764	55.3804 ug/L	99
48)	Methylcyclohexane	11.30	83	700951	57.4519 ug/L	98
49)	1,2-Dichloropropane	11.42	63	338396	50.4831 ug/L	89
50)	Bromodichloromethane	11.73	83	670981	56.9525 ug/L	100
51)	1,4-Dioxane	11.72	88	5139	152.6197 ug/L	86
52)	Dibromomethane	11.81	93	201226	50.7129 ug/L	96
53)	2-Chloroethyl Vinyl Ether	12.02	63	131607	45.6796 ug/L	98
54)	4-Methyl-2-Pentanone	12.05	58	55508	47.1653 ug/L	93
55)	cis-1,3-Dichloropropene	12.36	75	669136	55.7993 ug/L	100
56)	Dimethyl Disulfide	12.62	79	368346	57.5498 ug/L	98
59)	Toluene	12.78	91	1721616	50.6037 ug/L	100
60)	Ethyl Methacrylate	12.88	69	334116	51.1373 ug/L	96
62)	trans-1,3-Dichloropropene	12.96	75	598850	52.5249 ug/L	99
63)	1,1,2-Trichloroethane	13.18	97	257636	47.9905 ug/L	100
64)	2-Hexanone	13.12	58	50408	44.6823 ug/L #	96
65)	1,3-Dichloropropane	13.48	76	461776	48.3128 ug/L	98
66)	Tetrachloroethene	13.60	164	396562	54.4048 ug/L	96
67)	Dibromochloromethane	13.87	129	407562	51.9936 ug/L	100
68)	1,2-Dibromoethane	14.12	107	264335	49.4046 ug/L	100
69)	1-Chlorohexane	14.21	91	625960	53.9001 ug/L	98
70)	Chlorobenzene	14.62	112	1147614	50.0230 ug/L	99
71)	1,1,1,2-Tetrachloroethane	14.66	131	461856	51.9643 ug/L	98
72)	Ethylbenzene	14.66	106	624125	49.7229 ug/L	99
73)	m-,p-Xylene	14.75	106	1503763	102.1378 ug/L	99
74)	o-Xylene	15.31	106	737081	51.2070 ug/L	100
75)	Styrene	15.35	104	1225122	51.8929 ug/L	97
76)	Bromoform	15.83	173	248085	53.3581 ug/L	99
77)	Isopropylbenzene	15.73	105	1994935	52.1706 ug/L	99
79)	1,1,2,2-Tetrachloroethane	15.95	83	272075	43.5912 ug/L	99
81)	1,2,3-Trichloropropane	16.13	110	88111	45.8434 ug/L	85
82)	trans-1,4-Dichloro-2-Buten	16.18	53	93400	47.7038 ug/L	78
83)	n-Propylbenzene	16.24	91	2359407	48.6375 ug/L	99
84)	Bromobenzene	16.37	156	481990	48.3055 ug/L	97
85)	1,3,5-Trimethylbenzene	16.42	105	1760230	49.8132 ug/L	99
86)	2-Chlorotoluene	16.51	91	1680776	49.6264 ug/L	99
87)	4-Chlorotoluene	16.56	91	1429354	46.7216 ug/L	99
88)	a-Methylstyrene	16.83	118	848740	50.6254 ug/L	99
89)	tert-Butylbenzene	16.89	134	325684	47.1110 ug/L	98
90)	1,2,4-Trimethylbenzene	16.94	105	1777262	48.9126 ug/L	100
91)	sec-Butylbenzene	17.16	105	1996091	48.0164 ug/L	99
92)	p-Isopropyltoluene	17.32	119	1717647	48.6625 ug/L	99
93)	1,3-Dichlorobenzene	17.51	146	935639	48.6981 ug/L	98
94)	1,4-Dichlorobenzene	17.64	146	929018	47.9614 ug/L	97
95)	n-Butylbenzene	17.85	91	1670321	47.8670 ug/L	99
96)	1,2-Dichlorobenzene	18.14	146	798369	46.7212 ug/L	99
97)	1,2-Dibromo-3-Chloropropan	19.13	75	56411	45.0254 ug/L	99
98)	1,2,4-Trichlorobenzene	20.28	180	603545	47.2526 ug/L	99
99)	Hexachlorobutadiene	20.43	225	286706	48.0413 ug/L	98
100)	Naphthalene	20.65	128	820336	43.2287 ug/L	99
101)	1,2,3-Trichlorobenzene	20.96	180	475605	44.1363 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M413163.D 8260WT.M Mon Jun 27 12:27:21 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413163.D Vial: 3
Acq On : 22 Jun 2016 15:41 Operator: FJB
Sample : WG573770-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 16:05 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413163.D 8260WT.M Mon Jun 27 12:27:22 2016 Page 3

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413163.D Vial: 3
Acq On : 22 Jun 2016 15:41 Operator: FJB
Sample : WG573770-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	1.0000	1.0000	0.0	94	0.00
2	T	Dichlorodifluoromethane	0.4192	0.4520	-7.8	97	0.00
3	P	Chloromethane	0.2447	0.2430	0.7	96	0.00
4	C	Vinyl Chloride	0.2903	0.3342	-15.1	108	0.00
5	T	1,3-Butadiene	0.2434	0.3235	-32.9#	121	0.00
6	T	Bromomethane	0.1654	0.1269	23.3	71	0.00
7	T	Chloroethane	0.1597	0.1687	-5.6	101	0.00
8	T	Trichlorofluoromethane	0.5401	0.6548	-21.2	115	0.00
9	T	Diethyl ether	0.1290	0.1277	1.0	96	0.00
10	T	Isoprene	0.3280	0.3842	-17.1	113	0.00
11	T	Acrolein	0.0116	0.0087	24.7	72	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	0.2539	0.2790	-9.9	106	0.00
13	T	Acetone	0.0281	0.0329	-17.0	105	-0.01
14	C	1,1-Dichloroethene	0.4043	0.4598	-13.7	109	0.00
15	T	Tert-Butyl Alcohol	0.0090	0.0079	12.4	83	-0.01
16	T	Dimethyl Sulfide	0.1445	0.1455	-0.6	96	0.00
17	T	Iodomethane	0.0866	0.0983	-13.5	91	0.00
18	T	Methyl acetate	0.0760	0.0848	-11.6	108	0.00
19	T	Methylene Chloride	0.2573	0.2613	-1.6	98	0.00
20	T	Carbon Disulfide	0.8460	0.9895	-17.0	114	0.00
21	T	Acrylonitrile	0.0365	0.0378	-3.6	93	0.00
22	T	Methyl Tert Butyl Ether	0.5568	0.6006	-7.9	98	0.00
23	T	trans-1,2-Dichloroethene	0.3714	0.4035	-8.6	104	0.00
24	T	n-Hexane	0.2758	0.3161	-14.6	110	0.00
25	T	Diisopropyl ether	0.5789	0.5974	-3.2	98	0.00
26	T	Vinyl Acetate	0.2651	0.2303	13.1	79	0.00
27	P	1,1-Dichloroethane	0.4612	0.5090	-10.4	103	0.00
28	T	Ethyl-Tert-Butyl ether	0.6203	0.6799	-9.6	102	0.00
29	T	2-Butanone	0.0413	0.0399	3.5	88	0.00
30	T	Propionitrile	0.0120	0.0116	3.7	89	0.00
31	T	2,2-Dichloropropane	0.4860	0.6188	-27.3#	118	0.00
32	T	cis-1,2-Dichloroethene	0.2870	0.3062	-6.7	102	0.00
33	C	Chloroform	0.5446	0.6070	-11.5	105	0.00
34		1-Bromopropane	0.0411	0.0523	-27.2#	121	0.00
35	T	Bromochloromethane	0.1472	0.1603	-8.9	102	0.00
36	T	Tetrahydrofuran	0.0272	0.0247	9.1	91	0.00
37	S	Dibromofluoromethane	0.2686	0.2918	-8.6	101	0.00
38	T	1,1,1-Trichloroethane	0.5173	0.6175	-19.4	111	0.00
39	T	Cyclohexane	0.3581	0.3960	-10.6	108	0.00
40	T	1,1-Dichloropropene	0.3943	0.4449	-12.8	106	0.00
41	T	Tert-Amyl-Methyl ether	0.5634	0.6138	-9.0	102	0.00
42	T	Carbon Tetrachloride	0.4593	0.5596	-21.8	113	0.00
43	S	1,2-Dichloroethane-d4	0.2880	0.3258	-13.1	104	0.00
44		Heptane	0.0000	0.0000	0.0	0#	-2.54#
45	T	1,2-Dichloroethane	0.3505	0.4070	-16.1	106	0.00
46	T	Benzene	1.0269	1.0942	-6.5	100	0.00
47	T	Trichloroethene	0.2739	0.3034	-10.8	105	0.00
48	T	Methylcyclohexane	0.4161	0.4781	-14.9	110	0.00
49	C	1,2-Dichloropropane	0.2286	0.2308	-1.0	96	0.00
50	T	Bromodichloromethane	0.4018	0.4577	-13.9	102	0.00
51	T	1,4-Dioxane	0.0012	0.0009	23.5	68	0.00
52	T	Dibromomethane	0.1353	0.1373	-1.4	96	0.00
53	T	2-Chloroethyl Vinyl Ether	0.0983	0.0898	8.6	81	0.00
54	T	4-Methyl-2-Pentanone	0.0401	0.0379	5.7	85	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413163.D Vial: 3
Acq On : 22 Jun 2016 15:41 Operator: FJB
Sample : WG573770-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	0.4090	0.4564	-11.6	100	0.00
56	T	Dimethyl Disulfide	0.2183	0.2512	-15.1	107	0.00
57	I	Chlorobenzene-d5	1.0000	1.0000	0.0	102	0.00
58	S	Toluene-d8	1.3084	1.3810	-5.5	105	0.00
59	C	Toluene	1.5520	1.5708	-1.2	103	0.00
60	T	Ethyl Methacrylate	0.2981	0.3048	-2.3	99	0.00
61		Paraldehyde	0.0000	0.0000	0.0	0#	-13.14#
62	T	trans-1,3-Dichloropropene	0.5201	0.5464	-5.0	101	0.00
63	T	1,1,2-Trichloroethane	0.2449	0.2351	4.0	96	0.00
64	T	2-Hexanone	0.0515	0.0460	10.6	89	0.00
65	T	1,3-Dichloropropane	0.4360	0.4213	3.4	94	0.00
66	T	Tetrachloroethene	0.3325	0.3618	-8.8	113	0.00
67	T	Dibromochloromethane	0.3576	0.3719	-4.0	102	0.00
68	T	1,2-Dibromoethane	0.2441	0.2412	1.2	97	0.00
69	T	1-Chlorohexane	0.5298	0.5711	-7.8	112	0.00
70	P	Chlorobenzene	1.0466	1.0471	-0.0	103	0.00
71	T	1,1,1,2-Tetrachloroethane	0.4055	0.4214	-3.9	103	0.00
72	C	Ethylbenzene	0.5726	0.5694	0.6	103	0.00
73	T	m-,p-Xylene	0.6716	0.6860	-2.1	104	0.00
74	T	o-Xylene	0.6566	0.6725	-2.4	104	0.00
75	T	Styrene	1.0770	1.1178	-3.8	101	0.00
76	P	Bromoform	0.2121	0.2264	-6.7	104	0.00
77	T	Isopropylbenzene	1.7444	1.8201	-4.3	104	0.00
78	I	1,4-Dichlorobenzene-d4	1.0000	1.0000	0.0	108	0.00
79	P	1,1,2,2-Tetrachloroethane	0.5018	0.4375	12.8	92	0.00
80	S	p-Bromofluorobenzene	1.0352	1.0484	-1.3	111	0.00
81	T	1,2,3-Trichloropropane	0.1545	0.1417	8.3	99	0.00
82	T	trans-1,4-Dichloro-2-Butene	0.1376	0.1502	-9.2	108	0.00
83	T	n-Propylbenzene	3.9003	3.7940	2.7	104	0.00
84	T	Bromobenzene	0.8022	0.7751	3.4	105	0.00
85	T	1,3,5-Trimethylbenzene	2.8411	2.8305	0.4	106	0.00
86	T	2-Chlorotoluene	2.7231	2.7027	0.7	104	0.00
87	T	4-Chlorotoluene	2.4597	2.2984	6.6	107	0.00
88	T	a-Methylstyrene	1.3479	1.3648	-1.3	110	0.00
89	T	tert-Butylbenzene	0.5558	0.5237	5.8	106	0.00
90	T	1,2,4-Trimethylbenzene	2.9214	2.8579	2.2	106	0.00
91	T	sec-Butylbenzene	3.3424	3.2098	4.0	104	0.00
92	T	p-Isopropyltoluene	2.8379	2.7620	2.7	106	0.00
93	T	1,3-Dichlorobenzene	1.5448	1.5045	2.6	106	0.00
94	T	1,4-Dichlorobenzene	1.5574	1.4939	4.1	107	0.00
95	T	n-Butylbenzene	2.8056	2.6859	4.3	104	0.00
96	T	1,2-Dichlorobenzene	1.3739	1.2838	6.6	104	0.00
97	T	1,2-Dibromo-3-Chloropropane	0.1007	0.0907	9.9	95	0.00
98	T	1,2,4-Trichlorobenzene	1.0269	0.9705	5.5	103	0.00
99	T	Hexachlorobutadiene	0.4798	0.4610	3.9	108	0.00
100	T	Naphthalene	1.5257	1.3191	13.5	92	0.00
101	T	1,2,3-Trichlorobenzene	0.8664	0.7648	11.7	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M413163.D 8260WT.M Mon Jun 27 12:36:37 2016 Page 2

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413163.D Vial: 3
Acq On : 22 Jun 2016 15:41 Operator: FJB
Sample : WG573770-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	94	0.00
2	T	Dichlorodifluoromethane	50.0000	53.9222	-7.8	97	0.00
3	P	Chloromethane	50.0000	49.6540	0.7	96	0.00
4	C	Vinyl Chloride	50.0000	57.5583	-15.1	108	0.00
5	T	1,3-Butadiene	50.0000	66.4454	-32.9#	121	0.00
6	T	Bromomethane	50.0000	36.6673	26.7#	71	0.00
7	T	Chloroethane	50.0000	52.8222	-5.6	101	0.00
8	T	Trichlorofluoromethane	50.0000	60.6119	-21.2	115	0.00
9	T	Diethyl ether	100.0000	99.0505	0.9	96	0.00
10	T	Isoprene	50.0000	58.5676	-17.1	113	0.00
11	T	Acrolein	50.0000	37.6748	24.7	72	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	50.0000	54.9504	-9.9	106	0.00
13	T	Acetone	50.0000	58.5273	-17.1	105	-0.01
14	C	1,1-Dichloroethene	50.0000	56.8711	-13.7	109	0.00
15	T	Tert-Butyl Alcohol	200.0000	174.9747	12.5	83	-0.01
16	T	Dimethyl Sulfide	50.0000	50.3179	-0.6	96	0.00
17	T	Iodomethane	50.0000	44.1276	11.7	91	0.00
18	T	Methyl acetate	50.0000	55.7902	-11.6	108	0.00
19	T	Methylene Chloride	50.0000	50.7825	-1.6	98	0.00
20	T	Carbon Disulfide	50.0000	58.4851	-17.0	114	0.00
21	T	Acrylonitrile	50.0000	51.7966	-3.6	93	0.00
22	T	Methyl Tert Butyl Ether	50.0000	53.9358	-7.9	98	0.00
23	T	trans-1,2-Dichloroethene	50.0000	54.3180	-8.6	104	0.00
24	T	n-Hexane	50.0000	57.3174	-14.6	110	0.00
25	T	Diisopropyl ether	100.0000	103.1850	-3.2	98	0.00
26	T	Vinyl Acetate	50.0000	43.4350	13.1	79	0.00
27	P	1,1-Dichloroethane	50.0000	55.1763	-10.4	103	0.00
28	T	Ethyl-Tert-Butyl ether	100.0000	109.6148	-9.6	102	0.00
29	T	2-Butanone	50.0000	48.2362	3.5	88	0.00
30	T	Propionitrile	100.0000	96.2072	3.8	89	0.00
31	T	2,2-Dichloropropane	50.0000	63.6573	-27.3#	118	0.00
32	T	cis-1,2-Dichloroethene	50.0000	53.3528	-6.7	102	0.00
33	C	Chloroform	50.0000	55.7319	-11.5	105	0.00
34		1-Bromopropane	50.0000	63.5756	-27.2#	121	0.00
35	T	Bromochloromethane	50.0000	54.4744	-8.9	102	0.00
36	T	Tetrahydrofuran	100.0000	90.8490	9.2	91	0.00
37	S	Dibromofluoromethane	25.0000	27.1579	-8.6	101	0.00
38	T	1,1,1-Trichloroethane	50.0000	59.6822	-19.4	111	0.00
39	T	Cyclohexane	50.0000	55.2847	-10.6	108	0.00
40	T	1,1-Dichloropropene	50.0000	56.4182	-12.8	106	0.00
41	T	Tert-Amyl-Methyl ether	100.0000	108.9522	-9.0	102	0.00
42	T	Carbon Tetrachloride	50.0000	60.9134	-21.8	113	0.00
43	S	1,2-Dichloroethane-d4	25.0000	28.2849	-13.1	104	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	50.0000	58.0600	-16.1	106	0.00
46	T	Benzene	50.0000	53.2746	-6.5	100	0.00
47	T	Trichloroethene	50.0000	55.3804	-10.8	105	0.00
48	T	Methylcyclohexane	50.0000	57.4519	-14.9	110	0.00
49	C	1,2-Dichloropropane	50.0000	50.4831	-1.0	96	0.00
50	T	Bromodichloromethane	50.0000	56.9525	-13.9	102	0.00
51	T	1,4-Dioxane	200.0000	152.6197	23.7	68	0.00
52	T	Dibromomethane	50.0000	50.7129	-1.4	96	0.00
53	T	2-Chloroethyl Vinyl Ether	50.0000	45.6796	8.6	81	0.00
54	T	4-Methyl-2-Pentanone	50.0000	47.1653	5.7	85	0.00

(#) = Out of Range
8M413163.D 8260WT.M Mon Jun 27 12:36:39 2016 Page 1

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413163.D Vial: 3
Acq On : 22 Jun 2016 15:41 Operator: FJB
Sample : WG573770-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76578 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	50.0000	55.7993	-11.6	100	0.00
56	T	Dimethyl Disulfide	50.0000	57.5498	-15.1	107	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	102	0.00
58	S	Toluene-d8	25.0000	26.3867	-5.5	105	0.00
59	C	Toluene	50.0000	50.6037	-1.2	103	0.00
60	T	Ethyl Methacrylate	50.0000	51.1373	-2.3	99	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	50.0000	52.5249	-5.0	101	0.00
63	T	1,1,2-Trichloroethane	50.0000	47.9905	4.0	96	0.00
64	T	2-Hexanone	50.0000	44.6823	10.6	89	0.00
65	T	1,3-Dichloropropane	50.0000	48.3128	3.4	94	0.00
66	T	Tetrachloroethene	50.0000	54.4048	-8.8	113	0.00
67	T	Dibromochloromethane	50.0000	51.9936	-4.0	102	0.00
68	T	1,2-Dibromoethane	50.0000	49.4046	1.2	97	0.00
69	T	1-Chlorohexane	50.0000	53.9001	-7.8	112	0.00
70	P	Chlorobenzene	50.0000	50.0230	-0.0	103	0.00
71	T	1,1,1,2-Tetrachloroethane	50.0000	51.9643	-3.9	103	0.00
72	C	Ethylbenzene	50.0000	49.7229	0.6	103	0.00
73	T	m-,p-Xylene	100.0000	102.1378	-2.1	104	0.00
74	T	o-Xylene	50.0000	51.2070	-2.4	104	0.00
75	T	Styrene	50.0000	51.8929	-3.8	101	0.00
76	P	Bromoform	50.0000	53.3581	-6.7	104	0.00
77	T	Isopropylbenzene	50.0000	52.1706	-4.3	104	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	108	0.00
79	P	1,1,2,2-Tetrachloroethane	50.0000	43.5912	12.8	92	0.00
80	S	p-Bromofluorobenzene	25.0000	25.3202	-1.3	111	0.00
81	T	1,2,3-Trichloropropane	50.0000	45.8434	8.3	99	0.00
82	T	trans-1,4-Dichloro-2-Butene	50.0000	47.7038	4.6	108	0.00
83	T	n-Propylbenzene	50.0000	48.6375	2.7	104	0.00
84	T	Bromobenzene	50.0000	48.3055	3.4	105	0.00
85	T	1,3,5-Trimethylbenzene	50.0000	49.8132	0.4	106	0.00
86	T	2-Chlorotoluene	50.0000	49.6264	0.7	104	0.00
87	T	4-Chlorotoluene	50.0000	46.7216	6.6	107	0.00
88	T	a-Methylstyrene	50.0000	50.6254	-1.3	110	0.00
89	T	tert-Butylbenzene	50.0000	47.1110	5.8	106	0.00
90	T	1,2,4-Trimethylbenzene	50.0000	48.9126	2.2	106	0.00
91	T	sec-Butylbenzene	50.0000	48.0164	4.0	104	0.00
92	T	p-Isopropyltoluene	50.0000	48.6625	2.7	106	0.00
93	T	1,3-Dichlorobenzene	50.0000	48.6981	2.6	106	0.00
94	T	1,4-Dichlorobenzene	50.0000	47.9614	4.1	107	0.00
95	T	n-Butylbenzene	50.0000	47.8670	4.3	104	0.00
96	T	1,2-Dichlorobenzene	50.0000	46.7212	6.6	104	0.00
97	T	1,2-Dibromo-3-Chloropropane	50.0000	45.0254	9.9	95	0.00
98	T	1,2,4-Trichlorobenzene	50.0000	47.2526	5.5	103	0.00
99	T	Hexachlorobutadiene	50.0000	48.0413	3.9	108	0.00
100	T	Naphthalene	50.0000	43.2287	13.5	92	0.00
101	T	1,2,3-Trichlorobenzene	50.0000	44.1363	11.7	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M413163.D 8260WT.M Mon Jun 27 12:36:40 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413192.D Vial: 3
Acq On : 23 Jun 2016 21:00 Operator: FJB
Sample : WG573990-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76715 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 23 21:22:50 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	725028	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	565784	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	309614	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	215407	27.6524 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 110.60%		
43) 1,2-Dichloroethane-d4	10.30	65	235972	28.2513 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 113.00%		
58) Toluene-d8	12.68	98	743592	25.1124 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 100.44%		
80) p-Bromofluorobenzene	16.08	95	325934	25.4234 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 101.68%		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	636587	52.3691 ug/L	100
3) Chloromethane	3.62	50	336005	47.3540 ug/L	99
4) Vinyl Chloride	3.84	62	471297	55.9835 ug/L	99
5) 1,3-Butadiene	3.88	54	466509	66.0793 ug/L	94
6) Bromomethane	4.73	94	223807	44.2082 ug/L	96
7) Chloroethane	4.89	64	230630	49.7879 ug/L	97
8) Trichlorofluoromethane	5.37	101	914053	58.3521 ug/L	99
9) Diethyl ether	5.90	59	365966	97.8617 ug/L	97
10) Isoprene	5.94	67	528954	55.6063 ug/L	97
11) Acrolein	6.14	56	12820	38.2884 ug/L	100
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	398018	54.0552 ug/L	94
13) Acetone	6.23	43	45416	55.6757 ug/L	96
14) 1,1-Dichloroethene	6.47	61	635976	54.2465 ug/L	98
15) Tert-Butyl Alcohol	6.58	59	52882	202.2989 ug/L	98
16) Dimethyl Sulfide	6.73	62	202068	48.2045 ug/L	99
17) Iodomethane	6.99	142	143986	44.5471 ug/L	100
18) Methyl acetate	7.00	43	125266	56.8107 ug/L	98
19) Methylene Chloride	7.25	84	374398	50.1818 ug/L	100
20) Carbon Disulfide	7.29	76	1363583	55.5787 ug/L	100
21) Acrylonitrile	7.43	53	52604	49.7020 ug/L	93
22) Methyl Tert Butyl Ether	7.46	73	871485	53.9704 ug/L	99
23) trans-1,2-Dichloroethene	7.70	61	566696	52.6143 ug/L	100
24) n-Hexane	7.78	57	459061	57.3974 ug/L	99
25) Diisopropyl ether	8.12	45	1678736	99.9895 ug/L	96
26) Vinyl Acetate	8.30	43	314225	40.8778 ug/L	100
27) 1,1-Dichloroethane	8.33	63	710629	53.1263 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	1931789	107.3922 ug/L	98
29) 2-Butanone	8.88	43	57296	47.8066 ug/L	96
30) Propionitrile	8.98	54	34033	97.4427 ug/L	94
31) 2,2-Dichloropropane	9.10	77	868732	61.6292 ug/L	100
32) cis-1,2-Dichloroethene	9.17	96	432671	51.9888 ug/L	98
33) Chloroform	9.37	83	866433	54.8591 ug/L	100
34) 1-Bromopropane	9.51	122	73212	61.4154 ug/L	100
35) Bromochloromethane	9.61	130	227524	53.3132 ug/L	95
36) Tetrahydrofuran	9.63	42	69393	87.9098 ug/L	97
38) 1,1,1-Trichloroethane	9.91	97	883344	58.8810 ug/L	100
39) Cyclohexane	9.94	56	560595	53.9737 ug/L	97
40) 1,1-Dichloropropene	10.11	75	625468	54.6979 ug/L	100
41) Tert-Amyl-Methyl ether	10.21	73	1774358	108.6048 ug/L	100
42) Carbon Tetrachloride	10.25	117	798795	59.9616 ug/L	99

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413192.D Vial: 3
Acq On : 23 Jun 2016 21:00 Operator: FJB
Sample : WG573990-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76715 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 23 21:22:50 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
45)	1,2-Dichloroethane	10.42	62	592395	58.2757 ug/L		98
46)	Benzene	10.46	78	1539106	51.6786 ug/L		99
47)	Trichloroethene	11.21	130	432431	54.4401 ug/L		99
48)	Methylcyclohexane	11.30	83	686795	56.9140 ug/L		99
49)	1,2-Dichloropropane	11.43	63	327801	49.4432 ug/L		91
50)	Bromodichloromethane	11.73	83	655856	56.2842 ug/L		100
51)	1,4-Dioxane	11.72	88	6670	200.2779 ug/L		83
52)	Dibromomethane	11.81	93	198160	50.4924 ug/L		97
53)	2-Chloroethyl Vinyl Ether	12.02	63	128306	45.0262 ug/L		95
54)	4-Methyl-2-Pentanone	12.06	58	55902	48.0253 ug/L		91
55)	cis-1,3-Dichloropropene	12.36	75	645256	54.4029 ug/L		100
56)	Dimethyl Disulfide	12.63	79	358712	56.6642 ug/L		98
59)	Toluene	12.78	91	1659929	47.2580 ug/L		99
60)	Ethyl Methacrylate	12.88	69	327740	48.5859 ug/L		95
62)	trans-1,3-Dichloropropene	12.96	75	582725	49.5052 ug/L		99
63)	1,1,2-Trichloroethane	13.18	97	253367	45.7129 ug/L		100
64)	2-Hexanone	13.11	58	49571	42.5602 ug/L	#	95
65)	1,3-Dichloropropane	13.49	76	457785	46.3909 ug/L		96
66)	Tetrachloroethene	13.60	164	374094	49.7104 ug/L		98
67)	Dibromochloromethane	13.87	129	402599	49.7473 ug/L		98
68)	1,2-Dibromoethane	14.13	107	263257	47.6577 ug/L		98
69)	1-Chlorohexane	14.21	91	606471	50.5817 ug/L		98
70)	Chlorobenzene	14.62	112	1116015	47.1177 ug/L		98
71)	1,1,1,2-Tetrachloroethane	14.65	131	455262	49.6136 ug/L		99
72)	Ethylbenzene	14.65	106	606126	46.7722 ug/L		99
73)	m-,p-Xylene	14.75	106	1464859	96.3703 ug/L		98
74)	o-Xylene	15.31	106	700366	47.1280 ug/L		96
75)	Styrene	15.35	104	1184000	48.5758 ug/L		95
76)	Bromoform	15.83	173	242179	50.4518 ug/L		99
77)	Isopropylbenzene	15.73	105	1964495	49.7609 ug/L		99
79)	1,1,2,2-Tetrachloroethane	15.95	83	268196	43.1537 ug/L		99
81)	1,2,3-Trichloropropane	16.14	110	85820	44.8426 ug/L		85
82)	trans-1,4-Dichloro-2-Buten	16.18	53	93736	48.0756 ug/L		75
83)	n-Propylbenzene	16.24	91	2331548	48.2690 ug/L		99
84)	Bromobenzene	16.37	156	478746	48.1858 ug/L		95
85)	1,3,5-Trimethylbenzene	16.42	105	1722684	48.9595 ug/L		99
86)	2-Chlorotoluene	16.51	91	1605346	47.6022 ug/L		99
87)	4-Chlorotoluene	16.56	91	1451412	47.6458 ug/L		100
88)	a-Methylstyrene	16.83	118	820053	49.1238 ug/L		100
89)	tert-Butylbenzene	16.90	134	321775	46.7449 ug/L		97
90)	1,2,4-Trimethylbenzene	16.94	105	1755727	48.5269 ug/L		100
91)	sec-Butylbenzene	17.17	105	1988839	48.0468 ug/L		99
92)	p-Isopropyltoluene	17.32	119	1708694	48.6162 ug/L		99
93)	1,3-Dichlorobenzene	17.52	146	921602	48.1730 ug/L		98
94)	1,4-Dichlorobenzene	17.64	146	907047	47.0277 ug/L		99
95)	n-Butylbenzene	17.85	91	1655056	47.6327 ug/L		99
96)	1,2-Dichlorobenzene	18.14	146	790507	46.4592 ug/L		99
97)	1,2-Dibromo-3-Chloropropan	19.13	75	55858	44.7749 ug/L		98
98)	1,2,4-Trichlorobenzene	20.28	180	587639	46.2043 ug/L		98
99)	Hexachlorobutadiene	20.43	225	291731	49.0927 ug/L		98
100)	Naphthalene	20.65	128	821061	43.4522 ug/L		98
101)	1,2,3-Trichlorobenzene	20.96	180	466895	43.5136 ug/L		99

(#) = qualifier out of range (m) = manual integration
8M413192.D 8260WT.M Mon Jun 27 12:30:01 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413192.D Vial: 3
Acq On : 23 Jun 2016 21:00 Operator: FJB
Sample : WG573990-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76715 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 23 21:22 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413192.D 8260WT.M Mon Jun 27 12:30:03 2016 Page 3

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413192.D Vial: 3
Acq On : 23 Jun 2016 21:00 Operator: FJB
Sample : WG573990-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76715 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	1.0000	1.0000	0.0	93	0.00
2	T	Dichlorodifluoromethane	0.4192	0.4390	-4.7	93	0.00
3	P	Chloromethane	0.2447	0.2317	5.3	90	0.00
4	C	Vinyl Chloride	0.2903	0.3250	-12.0	104	0.00
5	T	1,3-Butadiene	0.2434	0.3217	-32.2#	119	0.00
6	T	Bromomethane	0.1654	0.1543	6.7	86	0.00
7	T	Chloroethane	0.1597	0.1591	0.4	94	0.00
8	T	Trichlorofluoromethane	0.5401	0.6304	-16.7	109	0.00
9	T	Diethyl ether	0.1290	0.1262	2.1	94	0.00
10	T	Isoprene	0.3280	0.3648	-11.2	106	0.00
11	T	Acrolein	0.0116	0.0088	23.5	73	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	0.2539	0.2745	-8.1	103	0.00
13	T	Acetone	0.0281	0.0313	-11.3	99	-0.01
14	C	1,1-Dichloroethene	0.4043	0.4386	-8.5	103	0.00
15	T	Tert-Butyl Alcohol	0.0090	0.0091	-1.2	95	0.00
16	T	Dimethyl Sulfide	0.1445	0.1394	3.6	91	0.00
17	T	Iodomethane	0.0866	0.0993	-14.7	91	0.00
18	T	Methyl acetate	0.0760	0.0864	-13.6	109	0.00
19	T	Methylene Chloride	0.2573	0.2582	-0.4	96	0.00
20	T	Carbon Disulfide	0.8460	0.9404	-11.2	107	0.00
21	T	Acrylonitrile	0.0365	0.0363	0.6	88	0.00
22	T	Methyl Tert Butyl Ether	0.5568	0.6010	-7.9	97	0.00
23	T	trans-1,2-Dichloroethene	0.3714	0.3908	-5.2	99	0.00
24	T	n-Hexane	0.2758	0.3166	-14.8	109	0.00
25	T	Diisopropyl ether	0.5789	0.5788	0.0	94	0.00
26	T	Vinyl Acetate	0.2651	0.2167	18.2	73	0.00
27	P	1,1-Dichloroethane	0.4612	0.4901	-6.3	98	0.00
28	T	Ethyl-Tert-Butyl ether	0.6203	0.6661	-7.4	99	0.00
29	T	2-Butanone	0.0413	0.0395	4.4	86	0.00
30	T	Propionitrile	0.0120	0.0117	2.5	89	0.00
31	T	2,2-Dichloropropane	0.4860	0.5991	-23.3	113	0.00
32	T	cis-1,2-Dichloroethene	0.2870	0.2984	-4.0	98	0.00
33	C	Chloroform	0.5446	0.5975	-9.7	103	0.00
34		1-Bromopropane	0.0411	0.0505	-22.8	115	0.00
35	T	Bromochloromethane	0.1472	0.1569	-6.6	99	0.00
36	T	Tetrahydrofuran	0.0272	0.0239	12.1	87	0.00
37	S	Dibromofluoromethane	0.2686	0.2971	-10.6	102	0.00
38	T	1,1,1-Trichloroethane	0.5173	0.6092	-17.8	108	0.00
39	T	Cyclohexane	0.3581	0.3866	-7.9	105	0.00
40	T	1,1-Dichloropropene	0.3943	0.4313	-9.4	102	0.00
41	T	Tert-Amyl-Methyl ether	0.5634	0.6118	-8.6	100	0.00
42	T	Carbon Tetrachloride	0.4593	0.5509	-19.9	110	0.00
43	S	1,2-Dichloroethane-d4	0.2880	0.3255	-13.0	103	0.00
44		Heptane	0.0000	0.0000	0.0	0#	-2.54#
45	T	1,2-Dichloroethane	0.3505	0.4085	-16.5	106	0.00
46	T	Benzene	1.0269	1.0614	-3.4	96	0.00
47	T	Trichloroethene	0.2739	0.2982	-8.9	102	0.00
48	T	Methylcyclohexane	0.4161	0.4736	-13.8	108	0.00
49	C	1,2-Dichloropropane	0.2286	0.2261	1.1	93	0.00
50	T	Bromodichloromethane	0.4018	0.4523	-12.6	100	0.00
51	T	1,4-Dioxane	0.0012	0.0012	0.0	88	0.00
52	T	Dibromomethane	0.1353	0.1367	-1.0	95	0.00
53	T	2-Chloroethyl Vinyl Ether	0.0983	0.0885	10.0	79	0.00
54	T	4-Methyl-2-Pentanone	0.0401	0.0386	4.0	86	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413192.D Vial: 3
Acq On : 23 Jun 2016 21:00 Operator: FJB
Sample : WG573990-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76715 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	0.4090	0.4450	-8.8	97	0.00
56	T	Dimethyl Disulfide	0.2183	0.2474	-13.3	104	0.00
57	I	Chlorobenzene-d5	1.0000	1.0000	0.0	105	0.00
58	S	Toluene-d8	1.3084	1.3143	-0.5	103	0.00
59	C	Toluene	1.5520	1.4669	5.5	99	0.00
60	T	Ethyl Methacrylate	0.2981	0.2896	2.8	97	0.00
61		Paraldehyde	0.0000	0.0000	0.0	0#	-13.14#
62	T	trans-1,3-Dichloropropene	0.5201	0.5150	1.0	98	0.00
63	T	1,1,2-Trichloroethane	0.2449	0.2239	8.6	94	0.00
64	T	2-Hexanone	0.0515	0.0438	14.9	87	0.00
65	T	1,3-Dichloropropane	0.4360	0.4046	7.2	94	0.00
66	T	Tetrachloroethene	0.3325	0.3306	0.6	107	0.00
67	T	Dibromochloromethane	0.3576	0.3558	0.5	100	0.00
68	T	1,2-Dibromoethane	0.2441	0.2327	4.7	96	0.00
69	T	1-Chlorohexane	0.5298	0.5360	-1.2	108	0.00
70	P	Chlorobenzene	1.0466	0.9863	5.8	100	0.00
71	T	1,1,1,2-Tetrachloroethane	0.4055	0.4023	0.8	102	0.00
72	C	Ethylbenzene	0.5726	0.5356	6.5	101	0.00
73	T	m-,p-Xylene	0.6716	0.6473	3.6	101	0.00
74	T	o-Xylene	0.6566	0.6189	5.7	99	0.00
75	T	Styrene	1.0770	1.0463	2.8	98	0.00
76	P	Bromoform	0.2121	0.2140	-0.9	102	0.00
77	T	Isopropylbenzene	1.7444	1.7361	0.5	102	0.00
78	I	1,4-Dichlorobenzene-d4	1.0000	1.0000	0.0	107	0.00
79	P	1,1,2,2-Tetrachloroethane	0.5018	0.4331	13.7	91	0.00
80	S	p-Bromofluorobenzene	1.0352	1.0527	-1.7	111	0.00
81	T	1,2,3-Trichloropropane	0.1545	0.1386	10.3	96	0.00
82	T	trans-1,4-Dichloro-2-Butene	0.1376	0.1514	-10.0	108	0.00
83	T	n-Propylbenzene	3.9003	3.7653	3.5	102	0.00
84	T	Bromobenzene	0.8022	0.7731	3.6	104	0.00
85	T	1,3,5-Trimethylbenzene	2.8411	2.7820	2.1	104	0.00
86	T	2-Chlorotoluene	2.7231	2.5925	4.8	100	0.00
87	T	4-Chlorotoluene	2.4597	2.3439	4.7	108	0.00
88	T	a-Methylstyrene	1.3479	1.3243	1.8	106	0.00
89	T	tert-Butylbenzene	0.5558	0.5196	6.5	105	0.00
90	T	1,2,4-Trimethylbenzene	2.9214	2.8354	2.9	105	0.00
91	T	sec-Butylbenzene	3.3424	3.2118	3.9	103	0.00
92	T	p-Isopropyltoluene	2.8379	2.7594	2.8	105	0.00
93	T	1,3-Dichlorobenzene	1.5448	1.4883	3.7	104	0.00
94	T	1,4-Dichlorobenzene	1.5574	1.4648	5.9	104	0.00
95	T	n-Butylbenzene	2.8056	2.6728	4.7	103	0.00
96	T	1,2-Dichlorobenzene	1.3739	1.2766	7.1	103	0.00
97	T	1,2-Dibromo-3-Chloropropane	0.1007	0.0902	10.4	94	0.00
98	T	1,2,4-Trichlorobenzene	1.0269	0.9490	7.6	101	0.00
99	T	Hexachlorobutadiene	0.4798	0.4711	1.8	110	0.00
100	T	Naphthalene	1.5257	1.3259	13.1	92	0.00
101	T	1,2,3-Trichlorobenzene	0.8664	0.7540	13.0	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M413192.D 8260WT.M Mon Jun 27 12:37:25 2016 Page 2

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413192.D Vial: 3
Acq On : 23 Jun 2016 21:00 Operator: FJB
Sample : WG573990-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76715 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	93	0.00
2	T	Dichlorodifluoromethane	50.0000	52.3691	-4.7	93	0.00
3	P	Chloromethane	50.0000	47.3540	5.3	90	0.00
4	C	Vinyl Chloride	50.0000	55.9835	-12.0	104	0.00
5	T	1,3-Butadiene	50.0000	66.0793	-32.2#	119	0.00
6	T	Bromomethane	50.0000	44.2082	11.6	86	0.00
7	T	Chloroethane	50.0000	49.7878	0.4	94	0.00
8	T	Trichlorofluoromethane	50.0000	58.3521	-16.7	109	0.00
9	T	Diethyl ether	100.0000	97.8617	2.1	94	0.00
10	T	Isoprene	50.0000	55.6063	-11.2	106	0.00
11	T	Acrolein	50.0000	38.2884	23.4	73	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	50.0000	54.0552	-8.1	103	0.00
13	T	Acetone	50.0000	55.6757	-11.4	99	-0.01
14	C	1,1-Dichloroethene	50.0000	54.2465	-8.5	103	0.00
15	T	Tert-Butyl Alcohol	200.0000	202.2989	-1.1	95	0.00
16	T	Dimethyl Sulfide	50.0000	48.2045	3.6	91	0.00
17	T	Iodomethane	50.0000	44.5471	10.9	91	0.00
18	T	Methyl acetate	50.0000	56.8107	-13.6	109	0.00
19	T	Methylene Chloride	50.0000	50.1818	-0.4	96	0.00
20	T	Carbon Disulfide	50.0000	55.5787	-11.2	107	0.00
21	T	Acrylonitrile	50.0000	49.7020	0.6	88	0.00
22	T	Methyl Tert Butyl Ether	50.0000	53.9704	-7.9	97	0.00
23	T	trans-1,2-Dichloroethene	50.0000	52.6143	-5.2	99	0.00
24	T	n-Hexane	50.0000	57.3974	-14.8	109	0.00
25	T	Diisopropyl ether	100.0000	99.9895	0.0	94	0.00
26	T	Vinyl Acetate	50.0000	40.8778	18.2	73	0.00
27	P	1,1-Dichloroethane	50.0000	53.1264	-6.3	98	0.00
28	T	Ethyl-Tert-Butyl ether	100.0000	107.3922	-7.4	99	0.00
29	T	2-Butanone	50.0000	47.8066	4.4	86	0.00
30	T	Propionitrile	100.0000	97.4427	2.6	89	0.00
31	T	2,2-Dichloropropane	50.0000	61.6292	-23.3	113	0.00
32	T	cis-1,2-Dichloroethene	50.0000	51.9888	-4.0	98	0.00
33	C	Chloroform	50.0000	54.8591	-9.7	103	0.00
34		1-Bromopropane	50.0000	61.4154	-22.8	115	0.00
35	T	Bromochloromethane	50.0000	53.3132	-6.6	99	0.00
36	T	Tetrahydrofuran	100.0000	87.9098	12.1	87	0.00
37	S	Dibromofluoromethane	25.0000	27.6524	-10.6	102	0.00
38	T	1,1,1-Trichloroethane	50.0000	58.8810	-17.8	108	0.00
39	T	Cyclohexane	50.0000	53.9737	-7.9	105	0.00
40	T	1,1-Dichloropropene	50.0000	54.6979	-9.4	102	0.00
41	T	Tert-Amyl-Methyl ether	100.0000	108.6048	-8.6	100	0.00
42	T	Carbon Tetrachloride	50.0000	59.9616	-19.9	110	0.00
43	S	1,2-Dichloroethane-d4	25.0000	28.2513	-13.0	103	0.00
44		Heptane	-1.0000	0.0000	0.0	0	-2.54#
45	T	1,2-Dichloroethane	50.0000	58.2756	-16.6	106	0.00
46	T	Benzene	50.0000	51.6786	-3.4	96	0.00
47	T	Trichloroethene	50.0000	54.4400	-8.9	102	0.00
48	T	Methylcyclohexane	50.0000	56.9140	-13.8	108	0.00
49	C	1,2-Dichloropropane	50.0000	49.4432	1.1	93	0.00
50	T	Bromodichloromethane	50.0000	56.2841	-12.6	100	0.00
51	T	1,4-Dioxane	200.0000	200.2779	-0.1	88	0.00
52	T	Dibromomethane	50.0000	50.4924	-1.0	95	0.00
53	T	2-Chloroethyl Vinyl Ether	50.0000	45.0262	9.9	79	0.00
54	T	4-Methyl-2-Pentanone	50.0000	48.0253	3.9	86	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413192.D Vial: 3
Acq On : 23 Jun 2016 21:00 Operator: FJB
Sample : WG573990-02 50ug/L CCV 8260 Inst : HPMS8
Misc : 1,1 STD76715 Multiplr: 1.00
MS Integration Params: RTEINT.P

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	cis-1,3-Dichloropropene	50.0000	54.4029	-8.8	97	0.00
56	T	Dimethyl Disulfide	50.0000	56.6642	-13.3	104	0.00
57	I	Chlorobenzene-d5	25.0000	25.0000	0.0	105	0.00
58	S	Toluene-d8	25.0000	25.1124	-0.4	103	0.00
59	C	Toluene	50.0000	47.2580	5.5	99	0.00
60	T	Ethyl Methacrylate	50.0000	48.5859	2.8	97	0.00
61		Paraldehyde	-1.0000	0.0000	0.0	0	-13.14#
62	T	trans-1,3-Dichloropropene	50.0000	49.5052	1.0	98	0.00
63	T	1,1,2-Trichloroethane	50.0000	45.7129	8.6	94	0.00
64	T	2-Hexanone	50.0000	42.5602	14.9	87	0.00
65	T	1,3-Dichloropropane	50.0000	46.3909	7.2	94	0.00
66	T	Tetrachloroethene	50.0000	49.7104	0.6	107	0.00
67	T	Dibromochloromethane	50.0000	49.7473	0.5	100	0.00
68	T	1,2-Dibromoethane	50.0000	47.6577	4.7	96	0.00
69	T	1-Chlorohexane	50.0000	50.5817	-1.2	108	0.00
70	P	Chlorobenzene	50.0000	47.1177	5.8	100	0.00
71	T	1,1,1,2-Tetrachloroethane	50.0000	49.6136	0.8	102	0.00
72	C	Ethylbenzene	50.0000	46.7722	6.5	101	0.00
73	T	m-,p-Xylene	100.0000	96.3703	3.6	101	0.00
74	T	o-Xylene	50.0000	47.1280	5.7	99	0.00
75	T	Styrene	50.0000	48.5758	2.8	98	0.00
76	P	Bromoform	50.0000	50.4518	-0.9	102	0.00
77	T	Isopropylbenzene	50.0000	49.7609	0.5	102	0.00
78	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	107	0.00
79	P	1,1,2,2-Tetrachloroethane	50.0000	43.1537	13.7	91	0.00
80	S	p-Bromofluorobenzene	25.0000	25.4234	-1.7	111	0.00
81	T	1,2,3-Trichloropropane	50.0000	44.8426	10.3	96	0.00
82	T	trans-1,4-Dichloro-2-Butene	50.0000	48.0756	3.8	108	0.00
83	T	n-Propylbenzene	50.0000	48.2690	3.5	102	0.00
84	T	Bromobenzene	50.0000	48.1858	3.6	104	0.00
85	T	1,3,5-Trimethylbenzene	50.0000	48.9595	2.1	104	0.00
86	T	2-Chlorotoluene	50.0000	47.6022	4.8	100	0.00
87	T	4-Chlorotoluene	50.0000	47.6458	4.7	108	0.00
88	T	a-Methylstyrene	50.0000	49.1238	1.8	106	0.00
89	T	tert-Butylbenzene	50.0000	46.7449	6.5	105	0.00
90	T	1,2,4-Trimethylbenzene	50.0000	48.5269	2.9	105	0.00
91	T	sec-Butylbenzene	50.0000	48.0468	3.9	103	0.00
92	T	p-Isopropyltoluene	50.0000	48.6162	2.8	105	0.00
93	T	1,3-Dichlorobenzene	50.0000	48.1730	3.7	104	0.00
94	T	1,4-Dichlorobenzene	50.0000	47.0277	5.9	104	0.00
95	T	n-Butylbenzene	50.0000	47.6327	4.7	103	0.00
96	T	1,2-Dichlorobenzene	50.0000	46.4592	7.1	103	0.00
97	T	1,2-Dibromo-3-Chloropropane	50.0000	44.7749	10.5	94	0.00
98	T	1,2,4-Trichlorobenzene	50.0000	46.2043	7.6	101	0.00
99	T	Hexachlorobutadiene	50.0000	49.0927	1.8	110	0.00
100	T	Naphthalene	50.0000	43.4522	13.1	92	0.00
101	T	1,2,3-Trichlorobenzene	50.0000	43.5136	13.0	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
8M413192.D 8260WT.M Mon Jun 27 12:37:29 2016 Page 2

2.1.1.5 Raw QC Data

BFB

Data File : C:\MSDCHEM\2\DATA\051316\8M412218.D Vial: 3
Acq On : 13 May 2016 12:36 Operator: TMB
Sample : WG568561-01 50ng BFB STD 8260 Inst : HPMS8
Misc : 1,1 STD76034 Multiplr: 1.00
MS Integration Params: rteint.p
Method : C:\MSDCHEM\2\METHODS\BFB.M (RTE Integrator)
Title : SOP:OVL MSV01





AutoFind: Scans 102, 103, 104; Background Corrected with Scan 97

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	6575	PASS
75	95	30	60	51.3	20171	PASS
95	95	100	100	100.0	39312	PASS
96	95	5	9	6.7	2649	PASS
173	174	0.00	2	0.4	122	PASS
174	95	50	100	85.5	33616	PASS
175	174	5	9	8.2	2746	PASS
176	174	95	101	99.2	33360	PASS
177	176	5	9	6.8	2267	PASS

8M412218.D BFB.M Fri May 13 12:47:24 2016

BFB

```
Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413124.D Vial: 1
Acq On    : 21 Jun 2016  17:23                        Operator: FJB
Sample    : WG573644-01 BFB 50ng 8260                 Inst    : HPMS8
Misc      : 1,1 STD76496                              Multiplr: 1.00
MS Integration Params: RTEINT.P
Method    : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title     : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
```



AutoFind: Scans 101, 102, 103; Background Corrected with Scan 96

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	10071	PASS
75	95	30	60	53.8	31493	PASS
95	95	100	100	100.0	58490	PASS
96	95	5	9	6.5	3806	PASS
173	174	0.00	2	0.4	174	PASS
174	95	50	100	81.7	47778	PASS
175	174	5	9	7.6	3655	PASS
176	174	95	101	96.2	45946	PASS
177	176	5	9	6.5	3004	PASS

8M413124.D 8260WT.M Fri Jun 24 17:18:42 2016

BFB

```
Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413161.D Vial: 1
Acq On    : 22 Jun 2016  14:38                        Operator: FJB
Sample    : WG573770-01 BFB 50ng 8260                 Inst    : HPMS8
Misc      : 1,1 STD76496                              Multiplr: 1.00
MS Integration Params: RTEINT.P
Method    : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title     : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
```



AutoFind: Scans 101, 102, 103; Background Corrected with Scan 97

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	8091	PASS
75	95	30	60	56.5	23343	PASS
95	95	100	100	100.0	41328	PASS
96	95	5	9	6.4	2639	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	77.4	31974	PASS
175	174	5	9	8.2	2612	PASS
176	174	95	101	95.1	30422	PASS
177	176	5	9	7.1	2156	PASS

8M413161.D 8260WT.M Mon Jun 27 12:36:23 2016

BFB

```
Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413190.D Vial: 1
Acq On    : 23 Jun 2016  19:56                        Operator: FJB
Sample    : WG573990-01 BFB 50ng 8260                 Inst    : HPMS8
Misc      : 1,1 STD76496                              Multiplr: 1.00
MS Integration Params: RTEINT.P
Method    : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title     : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
```



AutoFind: Scans 101, 102, 103; Background Corrected with Scan 96

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	10373	PASS
75	95	30	60	58.1	31797	PASS
95	95	100	100	100.0	54685	PASS
96	95	5	9	7.0	3803	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	76.1	41629	PASS
175	174	5	9	7.2	2991	PASS
176	174	95	101	95.9	39941	PASS
177	176	5	9	6.8	2702	PASS

8M413190.D 8260WT.M Mon Jun 27 12:37:11 2016

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413128.D Vial: 5
Acq On : 21 Jun 2016 19:21 Operator: FJB
Sample : WG573645-01 BLANK 8260 Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 19:43:27 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	793415	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	605610	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	336359	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	237002	27.8022 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 111.20%		
43) 1,2-Dichloroethane-d4	10.30	65	251076	27.4686 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 109.88%		
58) Toluene-d8	12.68	98	831111	26.2223 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 104.88%		
80) p-Bromofluorobenzene	16.08	95	347372	24.9412 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 99.76%		
Target Compounds					Qvalue
36) Tetrahydrofuran	9.63	42	779	0.9018 ug/L #	80

(#) = qualifier out of range (m) = manual integration
8M413128.D 8260WT.M Fri Jun 24 17:20:01 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413128.D Vial: 5
Acq On : 21 Jun 2016 19:21 Operator: FJB
Sample : WG573645-01 BLANK 8260 Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 19:43 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413128.D 8260WT.M Fri Jun 24 17:20:01 2016 Page 2



#36
Tetrahydrofuran
Concen: 0.90 ug/L
RT: 9.63 min Scan# 711
Delta R.T. -0.00 min
Lab File: 8M413128.D
Acq: 21 Jun 2016 19:21

Tgt Ion: 42 Resp: 779

Ion	Ratio	Lower	Upper
42	100		
72	62.8	33.8	78.8
71	30.2	32.0	74.8#



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413128.D Vial: 5
Acq On : 21 Jun 2016 19:21 Operator: FJB
Sample : WG573645-01 BLANK 8260 Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 17:12:25 2016 Quant Results File: A9FOOWT.RES

Quant Method : K:\ORGANICS\V...\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 10:02:03 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	793415	25.00 ug/L	0.00
11) Chlorobenzene-d5	14.57	117	605610	25.00 ug/L	0.00
12) 1,4-Dichlorobenzene-d4	17.60	152	336359	25.00 ug/L	0.00

Target Compounds					Qvalue
13) Cyclohexanone	15.83	55	886	6.4403 ug/L #	43

(#) = qualifier out of range (m) = manual integration
8M413128.D A9FOOWT.M Fri Jun 24 17:12:27 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413128.D Vial: 5
Acq On : 21 Jun 2016 19:21 Operator: FJB
Sample : WG573645-01 BLANK 8260 Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 17:12 2016 Quant Results File: A9FOOWT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\A9FOOWT.M (RTE Integrator)
Title : A9-FOO Water SOP:MSV01 05-17-16 HPMS8
Last Update : Thu May 19 10:02:03 2016
Response via : Initial Calibration



8M413128.D A9FOOWT.M Fri Jun 24 17:12:27 2016 Page 2



#13
Cyclohexanone
Concen: 6.44 ug/L
RT: 15.83 min Scan# 1311
Delta R.T. -0.00 min
Lab File: 8M413128.D
Acq: 21 Jun 2016 19:21

Tgt Ion: 55 Resp: 886

Ion	Ratio	Lower	Upper
55	100		
98	0.0	19.2	44.8#



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413165.D Vial: 5
Acq On : 22 Jun 2016 16:39 Operator: FJB
Sample : WG573771-01 BLANK 8260 Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 17:01:46 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	723354	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	554912	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	305326	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	207968	26.7592 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 107.04%		
43) 1,2-Dichloroethane-d4	10.30	65	231478	27.7774 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 111.12%		
58) Toluene-d8	12.68	98	755342	26.0090 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 104.04%		
80) p-Bromofluorobenzene	16.08	95	311112	24.6081 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 98.44%		
Target Compounds					Qvalue
36) Tetrahydrofuran	9.64	42	677	0.8596 ug/L #	24

(#) = qualifier out of range (m) = manual integration
8M413165.D 8260WT.M Mon Jun 27 12:39:56 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413165.D Vial: 5
Acq On : 22 Jun 2016 16:39 Operator: FJB
Sample : WG573771-01 BLANK 8260 Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 17:01 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413165.D 8260WT.M Mon Jun 27 12:39:57 2016 Page 2



#36
Tetrahydrofuran
Concen: 0.86 ug/L
RT: 9.64 min Scan# 712
Delta R.T. 0.01 min
Lab File: 8M413165.D
Acq: 22 Jun 2016 16:39

Tgt Ion: 42 Resp: 677

Ion	Ratio	Lower	Upper
42	100		
72	0.0	33.8	78.8#
71	0.0	32.0	74.8#



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413194.D Vial: 5
Acq On : 23 Jun 2016 21:59 Operator: FJB
Sample : WG573991-01 BLANK 8260 Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 23 22:21:08 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	695599	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	547838	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	300577	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	210658	28.1869 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 112.76%		
43) 1,2-Dichloroethane-d4	10.30	65	230039	28.7061 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 114.84%		
58) Toluene-d8	12.68	98	731476	25.5125 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 102.04%		
80) p-Bromofluorobenzene	16.08	95	315963	25.3867 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 101.56%		
Target Compounds					Qvalue
36) Tetrahydrofuran	9.62	42	624	0.8240 ug/L	82

(#) = qualifier out of range (m) = manual integration
8M413194.D 8260WT.M Mon Jun 27 12:41:31 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413194.D Vial: 5
Acq On : 23 Jun 2016 21:59 Operator: FJB
Sample : WG573991-01 BLANK 8260 Inst : HPMS8
Misc : 1,1 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 23 22:21 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413194.D 8260WT.M Mon Jun 27 12:41:32 2016 Page 2



#36
Tetrahydrofuran
Concen: 0.82 ug/L
RT: 9.62 min Scan# 710
Delta R.T. -0.01 min
Lab File: 8M413194.D
Acq: 23 Jun 2016 21:59

Tgt Ion: 42 Resp: 624

Ion	Ratio	Lower	Upper
42	100		
72	65.5	33.8	78.8
71	70.5	32.0	74.8



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413129.D Vial: 6
Acq On : 21 Jun 2016 19:50 Operator: FJB
Sample : WG573645-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 20:12:25 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	813093	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	608259	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	338876	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	234764	26.8732 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 107.48%		
43) 1,2-Dichloroethane-d4	10.30	65	255637	27.2908 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 109.16%		
58) Toluene-d8	12.68	98	823364	25.8647 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 103.44%		
80) p-Bromofluorobenzene	16.08	95	343893	24.5080 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 98.04%		
Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	385920	28.3093 ug/L	99
3) Chloromethane	3.62	50	189733	23.8434 ug/L	100
4) Vinyl Chloride	3.84	62	257972	27.3245 ug/L	99
5) 1,3-Butadiene	3.89	54	170429	21.5260 ug/L	98
6) Bromomethane	4.73	94	103832	18.9505 ug/L	98
7) Chloroethane	4.89	64	126869	24.4218 ug/L	97
8) Trichlorofluoromethane	5.37	101	429622	24.4560 ug/L	98
9) Diethyl ether	5.90	59	445162	106.1463 ug/L	98
10) Isoprene	5.94	67	269952	25.3050 ug/L	98
11) Acrolein	6.14	56	34375	91.5454 ug/L	96
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	196953	23.8513 ug/L	92
13) Acetone	6.23	43	18414	20.1289 ug/L	99
14) 1,1-Dichloroethene	6.47	61	292924	22.2792 ug/L	98
15) Tert-Butyl Alcohol	6.57	59	41436	141.3442 ug/L	99
16) Dimethyl Sulfide	6.73	62	182770	38.8785 ug/L	100
17) Iodomethane	6.99	142	54446	17.0625 ug/L	97
18) Methyl acetate	7.00	43	45533	18.4135 ug/L	99
19) Methylene Chloride	7.24	84	182767	21.8436 ug/L	97
20) Carbon Disulfide	7.29	76	600770	21.8348 ug/L	100
21) Acrylonitrile	7.43	53	24317	20.4871 ug/L	95
22) Methyl Tert Butyl Ether	7.46	73	413774	22.8493 ug/L	99
23) trans-1,2-Dichloroethene	7.70	61	273761	22.6642 ug/L	100
24) n-Hexane	7.78	57	207172	23.0976 ug/L	99
25) Diisopropyl ether	8.12	45	2021506	107.3647 ug/L	98
26) Vinyl Acetate	8.30	43	201601	23.3859 ug/L	99
27) 1,1-Dichloroethane	8.32	63	327791	21.8514 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	2130780	105.6249 ug/L	99
29) 2-Butanone	8.88	43	27143	20.1947 ug/L	94
30) Propionitrile	8.98	54	34357	87.7160 ug/L	94
31) 2,2-Dichloropropane	9.09	77	403710	25.5379 ug/L	100
32) cis-1,2-Dichloroethene	9.17	96	210713	22.5766 ug/L	100
33) Chloroform	9.37	83	399749	22.5692 ug/L	99
34) 1-Bromopropane	9.51	122	42874	32.0703 ug/L	98
35) Bromochloromethane	9.61	130	106778	22.3102 ug/L	95
36) Tetrahydrofuran	9.63	42	79690	90.0202 ug/L	99
38) 1,1,1-Trichloroethane	9.91	97	396102	23.5433 ug/L	98
39) Cyclohexane	9.94	56	263296	22.6043 ug/L	99
40) 1,1-Dichloropropene	10.11	75	291240	22.7108 ug/L	99
41) Tert-Amyl-Methyl ether	10.21	73	2008618	109.6275 ug/L	100
42) Carbon Tetrachloride	10.25	117	356159	23.8394 ug/L	98

(#) = qualifier out of range (m) = manual integration
8M413129.D 8260WT.M Fri Jun 24 17:14:48 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413129.D Vial: 6
Acq On : 21 Jun 2016 19:50 Operator: FJB
Sample : WG573645-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 20:12:25 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit	Qvalue
45)	1,2-Dichloroethane	10.42	62	272087	23.8670 ug/L	99
46)	Benzene	10.46	78	755460	22.6187 ug/L	99
47)	Trichloroethene	11.21	130	209527	23.5210 ug/L	99
48)	Methylcyclohexane	11.30	83	314343	23.2279 ug/L	99
49)	1,2-Dichloropropane	11.43	63	163555	21.9975 ug/L	96
50)	Bromodichloromethane	11.73	83	298750	22.8613 ug/L	100
51)	1,4-Dioxane	11.71	88	2412	64.5802 ug/L	85
52)	Dibromomethane	11.81	93	88284	20.0589 ug/L	96
53)	2-Chloroethyl Vinyl Ether	12.03	63	64637	20.2262 ug/L	96
54)	4-Methyl-2-Pentanone	12.05	58	28023	21.4670 ug/L	98
55)	cis-1,3-Dichloropropene	12.36	75	326532	24.5488 ug/L	97
56)	Dimethyl Disulfide	12.62	79	158616	22.3421 ug/L	94
59)	Toluene	12.78	91	794671	21.0444 ug/L	100
60)	Ethyl Methacrylate	12.88	69	153222	21.1283 ug/L	97
62)	trans-1,3-Dichloropropene	12.96	75	272048	21.4979 ug/L	99
63)	1,1,2-Trichloroethane	13.18	97	120470	20.2176 ug/L	99
64)	2-Hexanone	13.11	58	21435	17.1184 ug/L #	94
65)	1,3-Dichloropropane	13.48	76	228486	21.5374 ug/L	97
66)	Tetrachloroethene	13.60	164	174802	21.6060 ug/L	98
67)	Dibromochloromethane	13.87	129	179224	20.5994 ug/L	100
68)	1,2-Dibromoethane	14.13	107	120622	20.3115 ug/L	99
69)	1-Chlorohexane	14.21	91	272223	21.1188 ug/L	98
70)	Chlorobenzene	14.62	112	542483	21.3040 ug/L	100
71)	1,1,1,2-Tetrachloroethane	14.65	131	206850	20.9680 ug/L	97
72)	Ethylbenzene	14.65	106	278765	20.0090 ug/L	96
73)	m-,p-Xylene	14.75	106	703885	43.0736 ug/L	97
74)	o-Xylene	15.31	106	342035	21.4085 ug/L	99
75)	Styrene	15.35	104	556990	21.2558 ug/L	98
76)	Bromoform	15.83	173	103480	20.0520 ug/L	99
77)	Isopropylbenzene	15.73	105	938778	22.1188 ug/L	99
79)	1,1,2,2-Tetrachloroethane	15.95	83	121621	17.8795 ug/L	99
81)	1,2,3-Trichloropropane	16.14	110	40186	19.1848 ug/L	81
82)	trans-1,4-Dichloro-2-Buten	16.18	53	33643	16.1778 ug/L	95
83)	n-Propylbenzene	16.24	91	1119012	21.1660 ug/L	99
84)	Bromobenzene	16.37	156	217365	19.9887 ug/L	96
85)	1,3,5-Trimethylbenzene	16.42	105	813102	21.1133 ug/L	100
86)	2-Chlorotoluene	16.51	91	772842	20.9377 ug/L	99
87)	4-Chlorotoluene	16.56	91	686310	20.5842 ug/L	100
88)	a-Methylstyrene	16.82	118	388387	21.2566 ug/L	100
89)	tert-Butylbenzene	16.90	134	158309	21.0120 ug/L	95
90)	1,2,4-Trimethylbenzene	16.94	105	794341	20.0591 ug/L	99
91)	sec-Butylbenzene	17.17	105	940451	20.7578 ug/L	99
92)	p-Isopropyltoluene	17.32	119	796285	20.6998 ug/L	99
93)	1,3-Dichlorobenzene	17.51	146	427880	20.4344 ug/L	97
94)	1,4-Dichlorobenzene	17.64	146	430799	20.4070 ug/L	98
95)	n-Butylbenzene	17.85	91	752865	19.7966 ug/L	99
96)	1,2-Dichlorobenzene	18.14	146	371132	19.9285 ug/L	99
97)	1,2-Dibromo-3-Chloropropan	19.13	75	24263	17.7694 ug/L	97
98)	1,2,4-Trichlorobenzene	20.28	180	273983	19.6823 ug/L	99
99)	Hexachlorobutadiene	20.43	225	138397	21.2785 ug/L	97
100)	Naphthalene	20.65	128	332733	16.0884 ug/L	99
101)	1,2,3-Trichlorobenzene	20.96	180	219276	18.6714 ug/L	96

(#) = qualifier out of range (m) = manual integration
8M413129.D 8260WT.M Fri Jun 24 17:14:48 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413129.D Vial: 6
Acq On : 21 Jun 2016 19:50 Operator: FJB
Sample : WG573645-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 20:12 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413166.D Vial: 6
Acq On : 22 Jun 2016 17:13 Operator: FJB
Sample : WG573771-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 17:35:52 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	717279	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	551241	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	303556	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	212482	27.5716 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 110.28%		
43) 1,2-Dichloroethane-d4	10.30	65	233548	28.2631 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 113.04%		
58) Toluene-d8	12.68	98	743229	25.7624 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 103.04%		
80) p-Bromofluorobenzene	16.08	95	322867	25.6868 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 102.76%		
Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	367471	30.5567 ug/L	100
3) Chloromethane	3.62	50	168107	23.9477 ug/L	99
4) Vinyl Chloride	3.84	62	237940	28.5693 ug/L	99
5) 1,3-Butadiene	3.89	54	145361	20.8123 ug/L	96
6) Bromomethane	4.73	94	87164	18.0686 ug/L	97
7) Chloroethane	4.89	64	112178	24.4783 ug/L	98
8) Trichlorofluoromethane	5.37	101	396918	25.6125 ug/L	100
9) Diethyl ether	5.90	59	391516	105.8250 ug/L	97
10) Isoprene	5.94	67	240447	25.5501 ug/L	98
11) Acrolein	6.14	56	28422	85.8027 ug/L	97
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	174066	23.8955 ug/L	94
13) Acetone	6.24	43	17983	22.2836 ug/L	98
14) 1,1-Dichloroethene	6.47	61	262370	22.6210 ug/L	98
15) Tert-Butyl Alcohol	6.57	59	39305	151.9848 ug/L	99
16) Dimethyl Sulfide	6.73	62	156204	37.6659 ug/L	97
17) Iodomethane	6.99	142	37210	13.8558 ug/L	94
18) Methyl acetate	7.00	43	41812	19.1674 ug/L	99
19) Methylene Chloride	7.25	84	159561	21.6175 ug/L	98
20) Carbon Disulfide	7.29	76	530568	21.8592 ug/L	99
21) Acrylonitrile	7.43	53	21810	20.8294 ug/L	94
22) Methyl Tert Butyl Ether	7.46	73	374130	23.4199 ug/L	100
23) trans-1,2-Dichloroethene	7.70	61	238512	22.3836 ug/L	98
24) n-Hexane	7.78	57	175503	22.1806 ug/L	99
25) Diisopropyl ether	8.12	45	1743957	104.9964 ug/L	96
26) Vinyl Acetate	8.29	43	167340	22.0046 ug/L	100
27) 1,1-Dichloroethane	8.32	63	288871	21.8292 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	1890531	106.2340 ug/L	98
29) 2-Butanone	8.88	43	24012	20.2516 ug/L	99
30) Propionitrile	8.99	54	33230	96.1714 ug/L	97
31) 2,2-Dichloropropane	9.10	77	369805	26.5179 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	185007	22.4702 ug/L	99
33) Chloroform	9.37	83	356893	22.8412 ug/L	100
34) 1-Bromopropane	9.51	122	38563	32.6989 ug/L	98
35) Bromochloromethane	9.61	130	90818	21.5103 ug/L	96
36) Tetrahydrofuran	9.63	42	66658	85.3573 ug/L	97
38) 1,1,1-Trichloroethane	9.91	97	365510	24.6270 ug/L	99
39) Cyclohexane	9.93	56	225900	21.9845 ug/L	99
40) 1,1-Dichloropropene	10.11	75	255902	22.6207 ug/L	100
41) Tert-Amyl-Methyl ether	10.21	73	1807896	111.8530 ug/L	100
42) Carbon Tetrachloride	10.25	117	323502	24.5460 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M413166.D 8260WT.M Mon Jun 27 12:27:37 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413166.D Vial: 6
Acq On : 22 Jun 2016 17:13 Operator: FJB
Sample : WG573771-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 17:35:52 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit	Qvalue
45)	1,2-Dichloroethane	10.42	62	248337	24.6936 ug/L	99
46)	Benzene	10.46	78	661098	22.4375 ug/L	99
47)	Trichloroethene	11.21	130	184798	23.5161 ug/L	99
48)	Methylcyclohexane	11.30	83	272829	22.8533 ug/L	99
49)	1,2-Dichloropropane	11.43	63	142472	21.7216 ug/L	90
50)	Bromodichloromethane	11.73	83	271759	23.5737 ug/L	99
51)	1,4-Dioxane	11.72	88	2920	88.6251 ug/L	89
52)	Dibromomethane	11.81	93	80951	20.8496 ug/L	99
53)	2-Chloroethyl Vinyl Ether	12.02	63	55876	19.8203 ug/L	96
54)	4-Methyl-2-Pentanone	12.05	58	23648	20.5354 ug/L	96
55)	cis-1,3-Dichloropropene	12.36	75	290707	24.7749 ug/L	99
56)	Dimethyl Disulfide	12.63	79	138771	22.1579 ug/L	93
59)	Toluene	12.78	91	704595	20.5890 ug/L	99
60)	Ethyl Methacrylate	12.88	69	134164	20.4139 ug/L	100
62)	trans-1,3-Dichloropropene	12.96	75	248009	21.6254 ug/L	99
63)	1,1,2-Trichloroethane	13.18	97	107509	19.9087 ug/L	99
64)	2-Hexanone	13.11	58	19078	16.8120 ug/L #	88
65)	1,3-Dichloropropane	13.48	76	206231	21.4503 ug/L	97
66)	Tetrachloroethene	13.60	164	157755	21.5159 ug/L	96
67)	Dibromochloromethane	13.87	129	160851	20.4000 ug/L	100
68)	1,2-Dibromoethane	14.13	107	107833	20.0361 ug/L	98
69)	1-Chlorohexane	14.21	91	238474	20.4142 ug/L	99
70)	Chlorobenzene	14.62	112	475785	20.6174 ug/L	98
71)	1,1,1,2-Tetrachloroethane	14.65	131	187326	20.9530 ug/L	98
72)	Ethylbenzene	14.65	106	256653	20.3273 ug/L	98
73)	m-,p-Xylene	14.75	106	625804	42.2566 ug/L	94
74)	o-Xylene	15.31	106	304157	21.0069 ug/L	95
75)	Styrene	15.35	104	502553	21.1621 ug/L	96
76)	Bromoform	15.83	173	91459	19.5558 ug/L	99
77)	Isopropylbenzene	15.73	105	843877	21.9394 ug/L	100
79)	1,1,2,2-Tetrachloroethane	15.95	83	111182	18.2466 ug/L	99
81)	1,2,3-Trichloropropane	16.14	110	36636	19.5251 ug/L	78
82)	trans-1,4-Dichloro-2-Buten	16.18	53	31600	16.9336 ug/L	90
83)	n-Propylbenzene	16.24	91	1005131	21.2241 ug/L	99
84)	Bromobenzene	16.37	156	197914	20.3176 ug/L	95
85)	1,3,5-Trimethylbenzene	16.42	105	734640	21.2955 ug/L	99
86)	2-Chlorotoluene	16.51	91	676287	20.4537 ug/L	100
87)	4-Chlorotoluene	16.56	91	616450	20.6402 ug/L	99
88)	a-Methylstyrene	16.82	118	349834	21.3744 ug/L	98
89)	tert-Butylbenzene	16.89	134	137492	20.3723 ug/L	97
90)	1,2,4-Trimethylbenzene	16.94	105	722350	20.3636 ug/L	100
91)	sec-Butylbenzene	17.17	105	847665	20.8868 ug/L	100
92)	p-Isopropyltoluene	17.32	119	711448	20.6463 ug/L	100
93)	1,3-Dichlorobenzene	17.51	146	386130	20.5862 ug/L	99
94)	1,4-Dichlorobenzene	17.64	146	386877	20.4587 ug/L	100
95)	n-Butylbenzene	17.85	91	671169	19.7018 ug/L	100
96)	1,2-Dichlorobenzene	18.14	146	344448	20.6477 ug/L	98
97)	1,2-Dibromo-3-Chloropropan	19.13	75	21900	17.9051 ug/L	94
98)	1,2,4-Trichlorobenzene	20.28	180	240189	19.2623 ug/L	97
99)	Hexachlorobutadiene	20.43	225	124346	21.3426 ug/L	99
100)	Naphthalene	20.65	128	303563	16.3858 ug/L	99
101)	1,2,3-Trichlorobenzene	20.96	180	187462	17.8197 ug/L	99

(#) = qualifier out of range (m) = manual integration
8M413166.D 8260WT.M Mon Jun 27 12:27:37 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062216\8M413166.D Vial: 6
Acq On : 22 Jun 2016 17:13 Operator: FJB
Sample : WG573771-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 22 17:35 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



8M413166.D 8260WT.M Mon Jun 27 12:27:39 2016 Page 3

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413195.D Vial: 6
Acq On : 23 Jun 2016 22:28 Operator: FJB
Sample : WG573991-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 23 22:50:03 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	709070	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	541017	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	307359	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.66	111	209450	27.4928 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 109.96%		
43) 1,2-Dichloroethane-d4	10.31	65	229900	28.1438 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 112.56%		
58) Toluene-d8	12.68	98	729028	25.7477 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 103.00%		
80) p-Bromofluorobenzene	16.07	95	314680	24.7257 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 98.92%		
Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	339836	28.5859 ug/L	99
3) Chloromethane	3.62	50	159886	23.0403 ug/L	100
4) Vinyl Chloride	3.84	62	229600	27.8871 ug/L	99
5) 1,3-Butadiene	3.88	54	140587	20.3618 ug/L	99
6) Bromomethane	4.72	94	81952	17.2192 ug/L	98
7) Chloroethane	4.89	64	105860	23.3671 ug/L	98
8) Trichlorofluoromethane	5.38	101	375940	24.5397 ug/L	99
9) Diethyl ether	5.90	59	379335	103.7196 ug/L	97
10) Isoprene	5.94	67	220602	23.7127 ug/L	96
11) Acrolein	6.13	56	25939	79.2134 ug/L	100
12) 1,1,2-Trichloro-1,2,2-Trif	6.15	101	164574	22.8540 ug/L	93
13) Acetone	6.24	43	19152	24.0069 ug/L	87
14) 1,1-Dichloroethene	6.46	61	256541	22.3745 ug/L	96
15) Tert-Butyl Alcohol	6.57	59	45376	177.4915 ug/L	96
16) Dimethyl Sulfide	6.73	62	153966	37.5561 ug/L	98
17) Iodomethane	6.98	142	63771	21.9232 ug/L	96
18) Methyl acetate	7.00	43	43450	20.1489 ug/L	98
19) Methylene Chloride	7.25	84	155667	21.3341 ug/L	99
20) Carbon Disulfide	7.29	76	500362	20.8534 ug/L	99
21) Acrylonitrile	7.43	53	20108	19.4263 ug/L	95
22) Methyl Tert Butyl Ether	7.46	73	358831	22.7223 ug/L	99
23) trans-1,2-Dichloroethene	7.70	61	233011	22.1205 ug/L	99
24) n-Hexane	7.78	57	168096	21.4904 ug/L	97
25) Diisopropyl ether	8.12	45	1709292	104.1008 ug/L	97
26) Vinyl Acetate	8.30	43	136128	18.1075 ug/L	99
27) 1,1-Dichloroethane	8.32	63	280479	21.4404 ug/L	98
28) Ethyl-Tert-Butyl ether	8.70	59	1832290	104.1533 ug/L	98
29) 2-Butanone	8.88	43	24883	21.2291 ug/L	100
30) Propionitrile	8.98	54	32526	95.2238 ug/L	98
31) 2,2-Dichloropropane	9.10	77	352384	25.5613 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	179136	22.0090 ug/L	95
33) Chloroform	9.38	83	351389	22.7493 ug/L	99
34) 1-Bromopropane	9.51	122	39005	33.4565 ug/L	91
35) Bromochloromethane	9.60	130	90313	21.6383 ug/L	94
36) Tetrahydrofuran	9.63	42	64447	83.4814 ug/L	94
38) 1,1,1-Trichloroethane	9.90	97	361985	24.6718 ug/L	99
39) Cyclohexane	9.94	56	211867	20.8575 ug/L	96
40) 1,1-Dichloropropene	10.10	75	243522	21.7756 ug/L	100
41) Tert-Amyl-Methyl ether	10.21	73	1748040	109.4019 ug/L	100
42) Carbon Tetrachloride	10.24	117	317988	24.4070 ug/L	100

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413195.D Vial: 6
Acq On : 23 Jun 2016 22:28 Operator: FJB
Sample : WG573991-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 23 22:50:03 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
45)	1,2-Dichloroethane	10.42	62	245331	24.6771 ug/L		98
46)	Benzene	10.46	78	636186	21.8420 ug/L		99
47)	Trichloroethene	11.21	130	181612	23.3782 ug/L		99
48)	Methylcyclohexane	11.30	83	258623	21.9142 ug/L		99
49)	1,2-Dichloropropane	11.43	63	139989	21.5901 ug/L		90
50)	Bromodichloromethane	11.72	83	266207	23.3595 ug/L		100
51)	1,4-Dioxane	11.72	88	2551	78.3219 ug/L		94
52)	Dibromomethane	11.81	93	79612	20.7422 ug/L		96
53)	2-Chloroethyl Vinyl Ether	12.02	63	54338	19.4979 ug/L		95
54)	4-Methyl-2-Pentanone	12.06	58	22608	19.8596 ug/L		96
55)	cis-1,3-Dichloropropene	12.36	75	284141	24.4957 ug/L		100
56)	Dimethyl Disulfide	12.62	79	134379	21.7050 ug/L		88
59)	Toluene	12.79	91	690554	20.5600 ug/L		99
60)	Ethyl Methacrylate	12.88	69	136325	21.1347 ug/L		100
62)	trans-1,3-Dichloropropene	12.96	75	238780	21.2141 ug/L		100
63)	1,1,2-Trichloroethane	13.18	97	101653	19.1800 ug/L		99
64)	2-Hexanone	13.11	58	19262	17.2949 ug/L	#	96
65)	1,3-Dichloropropane	13.48	76	201584	21.3632 ug/L		96
66)	Tetrachloroethene	13.60	164	150481	20.9116 ug/L		99
67)	Dibromochloromethane	13.86	129	157618	20.3677 ug/L		100
68)	1,2-Dibromoethane	14.12	107	104511	19.7859 ug/L		99
69)	1-Chlorohexane	14.21	91	228001	19.8866 ug/L		98
70)	Chlorobenzene	14.63	112	469098	20.7118 ug/L		99
71)	1,1,1,2-Tetrachloroethane	14.66	131	183570	20.9209 ug/L		99
72)	Ethylbenzene	14.66	106	248486	20.0524 ug/L		96
73)	m-,p-Xylene	14.74	106	610436	41.9979 ug/L		96
74)	o-Xylene	15.31	106	301773	21.2361 ug/L		97
75)	Styrene	15.34	104	491214	21.0756 ug/L		97
76)	Bromoform	15.83	173	87737	19.1145 ug/L		98
77)	Isopropylbenzene	15.73	105	820778	21.7422 ug/L		100
79)	1,1,2,2-Tetrachloroethane	15.95	83	105306	17.0684 ug/L		100
81)	1,2,3-Trichloropropane	16.13	110	36317	19.1156 ug/L	#	31
82)	trans-1,4-Dichloro-2-Buten	16.19	53	29611	15.7173 ug/L	#	4
83)	n-Propylbenzene	16.24	91	1001969	20.8955 ug/L		99
84)	Bromobenzene	16.37	156	191590	19.4250 ug/L		97
85)	1,3,5-Trimethylbenzene	16.42	105	724823	20.7509 ug/L		98
86)	2-Chlorotoluene	16.51	91	616163	18.4047 ug/L		92
87)	4-Chlorotoluene	16.56	91	670668	22.1777 ug/L		91
88)	a-Methylstyrene	16.83	118	336395	20.2990 ug/L		98
89)	tert-Butylbenzene	16.89	134	136372	19.9564 ug/L		99
90)	1,2,4-Trimethylbenzene	16.94	105	714803	19.9015 ug/L		99
91)	sec-Butylbenzene	17.16	105	828940	20.1726 ug/L		99
92)	p-Isopropyltoluene	17.32	119	690041	19.7773 ug/L		99
93)	1,3-Dichlorobenzene	17.51	146	383600	20.1982 ug/L		98
94)	1,4-Dichlorobenzene	17.64	146	379062	19.7974 ug/L		100
95)	n-Butylbenzene	17.84	91	654053	18.9618 ug/L		99
96)	1,2-Dichlorobenzene	18.14	146	327323	19.3784 ug/L		99
97)	1,2-Dibromo-3-Chloropropan	19.13	75	21199	17.1175 ug/L		94
98)	1,2,4-Trichlorobenzene	20.27	180	233503	18.4944 ug/L		98
99)	Hexachlorobutadiene	20.43	225	122070	20.6927 ug/L		98
100)	Naphthalene	20.64	128	285406	15.2151 ug/L		98
101)	1,2,3-Trichlorobenzene	20.95	180	185957	17.4579 ug/L		97

(#) = qualifier out of range (m) = manual integration
8M413195.D 8260WT.M Mon Jun 27 12:30:19 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062316\8M413195.D Vial: 6
Acq On : 23 Jun 2016 22:28 Operator: FJB
Sample : WG573991-02 20ug/L LCS 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 23 22:50 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413130.D Vial: 7
Acq On : 21 Jun 2016 20:19 Operator: FJB
Sample : WG573645-03 20ug/L LCS2 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 20:41:24 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.71	96	801570	25.00 ug/L	0.00
57) Chlorobenzene-d5	14.57	117	606723	25.00 ug/L	0.00
78) 1,4-Dichlorobenzene-d4	17.60	152	337211	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.66	111	232312	26.9748 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 107.88%		
43) 1,2-Dichloroethane-d4	10.30	65	244814	26.5111 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 106.04%		
58) Toluene-d8	12.68	98	807456	25.4292 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 101.72%		
80) p-Bromofluorobenzene	16.08	95	345700	24.7584 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 99.04%		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.16	85	375654	27.9524 ug/L	99
3) Chloromethane	3.62	50	183223	23.3563 ug/L	99
4) Vinyl Chloride	3.84	62	251449	27.0165 ug/L	100
5) 1,3-Butadiene	3.88	54	171496	21.9722 ug/L	99
6) Bromomethane	4.73	94	99370	18.4182 ug/L	97
7) Chloroethane	4.89	64	120768	23.5816 ug/L	97
8) Trichlorofluoromethane	5.37	101	405508	23.4152 ug/L	99
9) Diethyl ether	5.90	59	429224	103.8172 ug/L	98
10) Isoprene	5.93	67	259996	24.7221 ug/L	98
11) Acrolein	6.13	56	33670	90.9570 ug/L	94
12) 1,1,2-Trichloro-1,2,2-Trif	6.16	101	190538	23.4061 ug/L	94
13) Acetone	6.24	43	19487	21.6080 ug/L	98
14) 1,1-Dichloroethene	6.46	61	284861	21.9774 ug/L	97
15) Tert-Butyl Alcohol	6.57	59	44593	154.2999 ug/L	96
16) Dimethyl Sulfide	6.73	62	172245	37.1664 ug/L	97
17) Iodomethane	6.99	142	67479	20.7087 ug/L	100
18) Methyl acetate	6.99	43	45353	18.6044 ug/L	99
19) Methylene Chloride	7.24	84	175900	21.3251 ug/L	97
20) Carbon Disulfide	7.29	76	574700	21.1876 ug/L	100
21) Acrylonitrile	7.43	53	23931	20.4517 ug/L	98
22) Methyl Tert Butyl Ether	7.46	73	395219	22.1384 ug/L	100
23) trans-1,2-Dichloroethene	7.70	61	264912	22.2469 ug/L	99
24) n-Hexane	7.78	57	204188	23.0922 ug/L	98
25) Diisopropyl ether	8.12	45	1967896	106.0199 ug/L	98
26) Vinyl Acetate	8.30	43	186340	21.9263 ug/L	99
27) 1,1-Dichloroethane	8.32	63	312098	21.1043 ug/L	100
28) Ethyl-Tert-Butyl ether	8.70	59	2048753	103.0187 ug/L	99
29) 2-Butanone	8.88	43	26114	19.7084 ug/L	96
30) Propionitrile	8.98	54	34706	89.8808 ug/L	96
31) 2,2-Dichloropropane	9.09	77	386993	24.8323 ug/L	98
32) cis-1,2-Dichloroethene	9.17	96	200847	21.8289 ug/L	97
33) Chloroform	9.37	83	383139	21.9424 ug/L	100
34) 1-Bromopropane	9.51	122	41941	31.8234 ug/L	100
35) Bromochloromethane	9.60	130	100661	21.3345 ug/L	95
36) Tetrahydrofuran	9.63	42	73556	84.2855 ug/L	99
38) 1,1,1-Trichloroethane	9.91	97	386918	23.3280 ug/L	100
39) Cyclohexane	9.94	56	252690	22.0057 ug/L	100
40) 1,1-Dichloropropene	10.11	75	282605	22.3542 ug/L	99
41) Tert-Amyl-Methyl ether	10.21	73	1942419	107.5385 ug/L	99
42) Carbon Tetrachloride	10.24	117	339027	23.0189 ug/L	100

(#) = qualifier out of range (m) = manual integration
8M413130.D 8260WT.M Fri Jun 24 17:14:54 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413130.D Vial: 7
Acq On : 21 Jun 2016 20:19 Operator: FJB
Sample : WG573645-03 20ug/L LCS2 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 20:41:24 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\2\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc	Unit		Qvalue
45)	1,2-Dichloroethane	10.42	62	258437	22.9955	ug/L		99
46)	Benzene	10.46	78	728220	22.1166	ug/L		99
47)	Trichloroethene	11.21	130	203927	23.2215	ug/L		99
48)	Methylcyclohexane	11.30	83	305222	22.8782	ug/L		99
49)	1,2-Dichloropropane	11.43	63	160378	21.8803	ug/L		95
50)	Bromodichloromethane	11.73	83	286882	22.2687	ug/L		100
51)	1,4-Dioxane	11.72	88	4962	134.7651	ug/L		91
52)	Dibromomethane	11.81	93	86145	19.8542	ug/L		95
53)	2-Chloroethyl Vinyl Ether	12.02	63	61951	19.6644	ug/L		99
54)	4-Methyl-2-Pentanone	12.06	58	24431	18.9844	ug/L		93
55)	cis-1,3-Dichloropropene	12.36	75	314346	23.9724	ug/L		98
56)	Dimethyl Disulfide	12.62	79	151414	21.6343	ug/L		94
59)	Toluene	12.78	91	779721	20.7007	ug/L		98
60)	Ethyl Methacrylate	12.88	69	145866	20.1649	ug/L		98
62)	trans-1,3-Dichloropropene	12.96	75	259013	20.5196	ug/L		99
63)	1,1,2-Trichloroethane	13.18	97	114431	19.2527	ug/L		99
64)	2-Hexanone	13.11	58	22675	18.1545	ug/L	#	94
65)	1,3-Dichloropropane	13.48	76	220977	20.8823	ug/L		99
66)	Tetrachloroethene	13.60	164	173697	21.5238	ug/L		94
67)	Dibromochloromethane	13.87	129	173994	20.0489	ug/L		100
68)	1,2-Dibromoethane	14.12	107	116731	19.7061	ug/L		98
69)	1-Chlorohexane	14.21	91	262320	20.4021	ug/L		99
70)	Chlorobenzene	14.62	112	521869	20.5464	ug/L		99
71)	1,1,1,2-Tetrachloroethane	14.65	131	201136	20.4404	ug/L		99
72)	Ethylbenzene	14.65	106	275214	19.8041	ug/L		96
73)	m-,p-Xylene	14.75	106	679738	41.7013	ug/L		95
74)	o-Xylene	15.31	106	336370	21.1072	ug/L		99
75)	Styrene	15.35	104	545154	20.8568	ug/L		97
76)	Bromoform	15.83	173	98789	19.1915	ug/L		99
77)	Isopropylbenzene	15.73	105	925159	21.8531	ug/L		99
79)	1,1,2,2-Tetrachloroethane	15.95	83	117825	17.4069	ug/L		99
81)	1,2,3-Trichloropropane	16.14	110	40221	19.2963	ug/L		75
82)	trans-1,4-Dichloro-2-Buten	16.18	53	32499	15.7229	ug/L		94
83)	n-Propylbenzene	16.24	91	1101437	20.9364	ug/L		100
84)	Bromobenzene	16.37	156	214213	19.7961	ug/L		96
85)	1,3,5-Trimethylbenzene	16.42	105	798614	20.8395	ug/L		100
86)	2-Chlorotoluene	16.52	91	760317	20.7001	ug/L		99
87)	4-Chlorotoluene	16.56	91	661306	19.9322	ug/L		100
88)	a-Methylstyrene	16.83	118	382596	21.0431	ug/L		98
89)	tert-Butylbenzene	16.90	134	152707	20.3685	ug/L		97
90)	1,2,4-Trimethylbenzene	16.94	105	780583	19.8091	ug/L		99
91)	sec-Butylbenzene	17.16	105	925679	20.5326	ug/L		100
92)	p-Isopropyltoluene	17.32	119	770708	20.1338	ug/L		98
93)	1,3-Dichlorobenzene	17.52	146	423219	20.3116	ug/L		97
94)	1,4-Dichlorobenzene	17.64	146	424902	20.2270	ug/L		97
95)	n-Butylbenzene	17.85	91	740475	19.5669	ug/L		99
96)	1,2-Dichlorobenzene	18.14	146	360566	19.4567	ug/L		98
97)	1,2-Dibromo-3-Chloropropan	19.13	75	24053	17.7026	ug/L		99
98)	1,2,4-Trichlorobenzene	20.28	180	265360	19.1570	ug/L		98
99)	Hexachlorobutadiene	20.43	225	136867	21.1471	ug/L		100
100)	Naphthalene	20.65	128	338862	16.4656	ug/L		100
101)	1,2,3-Trichlorobenzene	20.96	180	216043	18.4869	ug/L		97

(#) = qualifier out of range (m) = manual integration
8M413130.D 8260WT.M Fri Jun 24 17:14:55 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : K:\ORGANICS\VOLATILE\HPMS8\DATA\062116\8M413130.D Vial: 7
Acq On : 21 Jun 2016 20:19 Operator: FJB
Sample : WG573645-03 20ug/L LCS2 8260 Inst : HPMS8
Misc : 1,1 STD76744 Multiplr: 1.00
MS Integration Params: RTEINT.P
Quant Time: Jun 21 20:41 2016 Quant Results File: 8260WT.RES

Method : K:\ORGANICS\VOLATILE\HPMS8\METHODS\8260WT.M (RTE Integrator)
Title : Method 8260B/624 WTR-SOP:OVLMSV01 05-13-16 HPMS 8
Last Update : Sat May 14 18:08:06 2016
Response via : Initial Calibration



2.2 Semivolatiles Data

2.2.1 Semivolatiles GC/MS Data (Dioxane)

2.2.1.1 Summary Data

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-01	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	18CPTMW23SW-061516	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 15:30
Collect Date:	06/15/2016 08:00	**Dilution:**	1	**File ID:**	15M18514
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	32.4	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-02	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	18CPTMW23-061516	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 15:53
Collect Date:	06/15/2016 09:05	**Dilution:**	1	**File ID:**	15M18515
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	7.62	J	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	133	20	129	*

J	The reported result is an estimated value.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-02	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	18CPTMW23-061516	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/25/2016 12:46
Collect Date:	06/15/2016 09:05	**Dilution:**	10	**File ID:**	15M18553
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	8.02	J	20.0	10.0	5.00

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	109	20	129	

J	Estimated value ; the analyte concentration was less than the LOQ.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-03	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	18CPTMW19SW-061516	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 16:15
Collect Date:	06/15/2016 10:15	**Dilution:**	1	**File ID:**	15M18516
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	37.5	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	18CPTMW19-061516	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 16:38
Collect Date:	06/15/2016 11:15	**Dilution:**	1	**File ID:**	15M18517
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	40.5	20	129	

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.2.1.2 QC Summary Data

Example 8270 Calculations

1.0 Calculating the Response Factor (RF) from the initial calibration (ICAL) data:

RF = [(Ax) (Cis)] / [(Ais) (Cx)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured:	1261197
Cis = Concentration of the specific internal standard (ug/mL)	40
Ais = Area of the characteristic ion of the specific internal standard	608044
Cx = Concentration of the compound in the standard being measured (ug/mL)	50
RF = Calculated Response Factor	**1.65935**

2.0 Calculating the concentration (C) of a compound in water using the data from the prep log and quantitation report: *

Cx = [(Ax) (Cis) (Vf) (D)] / [(Ais) (RF) (Vi)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured	367250
Cis = Concentration of the specific internal standard (ug/mL)	40
Vf = Final volume of sample extract from prep log (mL)	1
D = Dilution factor for sample as a multiplier (10x = 10)	1
Ais = Area of the characteristic ion of the specific internal standard	511641
RF = Average RF from the ICAL	1.65935
Vi = Initial volume of sample extracted from prep log (mL)	1021
Cx = Concentration of the compound in the sample being measured (ug/mL)	0.016947
Cx = Concentration of the compound in the sample being measured (ug/L)	**16.947**

3.0 Calculating the concentration (C) of a compound in soil using the data from the prep log and quantitation report: *

Cx = [(Ax) (Cis) (Vf) (D)] / [(Ais) (RF) (Wi)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured	367250
Cis = Concentration of the specific internal standard (ug/mL)	40
Vf = Final volume of sample extract from prep log (mL)	1
D = Dilution factor for sample as a multiplier (10x = 10)	1
Ais = Area of the characteristic ion of the specific internal standard	511641
RF = Average RF from the ICAL	1.65935
Wi = Initial weight of sample extracted (g) from prep log	30
Cx = Concentration of the compound in the sample being measured (ug/g)	0.576763
Cx = Concentration of the compound in the sample being measured (ug/kg)	**576.7627**

Dry weight correction:

Percent solids (PCT_S)	50
Cd = (Cx) (100)/PCT_S	**1153.525** ug/kg

* Concentrations appearing on the instrument quantitation reports are on-column results and do not take into account initial volume, final volume, and the dilution factor.

4.0 Concentration from Linear Regression

Step 1: Retrieve Curve Data From Plot, y = mx + b

y = response ratio = response of analyte / response of IS = Ax/Ais

x = amount ratio = concentration analyte/concentration internal standard = Cx / Cis

m = slope from curve plot

b = intercept from curve plot

Step 2: Calculate y from Quantitation Report

y = 16790/784838 = 0.02139

Step 3: Solve for x

x = (y - b)/m = [(0.02139 -(- 0.0435)]/ 0.0783 = 0.829

Step 4: Solve for analyte concentration Cx

Cx = Cis (x) = (25.0)(0.829) = 20.72 ug/L

Example Spreadsheet Calculation:

Slope from curve, m:	0.0783	
Intercept from curve, b:	-0.0435	
Area of analyte, Ax:	16790	
Area of Internal Standard , Ais:	784484	
Concentration of IS, Cis	25.00	ug/L
Response Ratio (y) :	0.021403	
Amount Ratio:	0.828897	
Concentration (Cx):	20.72241	ug/L

5.0 Concentration from Quadratic Regression

Step 1 - Retrieve Curve Data from Plot, y = Ax^2 + Bx + C
Where:
Ax^2 + Bx + (C - y) = 0
A, B, C = constants from the ICAL quadratic regression
y = Response ratio = Area of analyte/Area of internal standard (IS)
x = Amount ratio = Concentration of analyte/concentration of IS

Step 2: Calculate y from Quantitation Report

y = Ax/Ais

Step 3: Solve for x using the quadratic formula
Ax^2 + Bx + C - y = 0

$$x = \frac{b \pm \sqrt{\left(b^2 - 4a(c - y)\right)}}{2a}$$ (Two possible solutions)

Step 4: Solve for analyte concentration Cx

Cx = (Cis)(Amount ratio)

Example Spreadsheet Calculation:

Value of A from plot:	**0.0259**	
Value of B from plot:	**0.0596**	
Value of C from plot:	**-0.0165**	
Area of analyte from quantitation report:	**203233**	
Area of IS from quantiation report:	**1425653**	
Response ratio, y:	**0.142554**	
C - y:	**-0.15905**	
Root 1 - Computed amount ratio , X1:	**-3.88278**	
Root 2 - Computed amount ratio , X2:	**1.581623**	use this solution
Concentration of IS, Cis:	**40.00**	
Concentration of analyte, Cx:	**63.26**	ug/L

Microbac Laboratories Inc.
Continuous Sample Extract Log

Workgroup: WG573299
Analyst: CPD
Spike Analyst: CPD
Method: 3520C
Run Date: 06/21/2016 17:01
SOP: EXB01 Revison 20
Spike Witness: JLD
Surr Solution: STD76628

TIME ON: 17:45 OFF: 12:10 ON: 16:00 OFF: 11:05

CH2CL2 Lot #: COA18982
sodium sulfate Lot #: COA18814
10N NaOH Lot #: RGT37076
1:1 H2SO4 Lot #: RGT36755

	SAMPLE #	Type	Reference	Prod	pH	Init Amnt	Surr Amnt	Spike Amnt	Spike Sol	Final Vol	Color
1	L16060927-01	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
2	L16060927-02	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
3	L16060927-03	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
4	L16060927-04	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
5	L16060930-03	SAMP		827-DIOX	<2>12	1000 mL	.05 mL			1 mL	Colored
6	L16060978-03	SAMP		827-DIOX	<2>12	800 mL	.05 mL			1 mL	Colored
7	L16060978-04	SAMP		827-DIOX	<2>12	890 mL	.05 mL			1 mL	Colored
8	L16061026-04	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
9	L16061026-05	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
10	L16061026-07	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
11	L16061055-04	SAMP		827-DIOX	<2>12	520 mL	.05 mL			1 mL	Colored
12	L16061055-07	SAMP		827-DIOX	<2>12	880 mL	.05 mL			1 mL	Colored
13	WG573299-01	BLANK		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
14	WG573299-02	LCS		827-DIOXANE	<2>12	1000 mL	.05 mL	.1 mL	STD75281	1 mL	Transparent
15	WG573299-03	LCS2		827-DIOXANE	<2>12	1000 mL	.05 mL	.1 mL	STD75281	1 mL	Transparent

Due to insufficient sample volume, this preparation batch failed to include the method prescribed MS and MSD.

Analyst: Chelsie Davis

Reviewer: Justin Hesson

EXTRACT - Modified 01/23/2009
PDF ID: 4819876
Report generated: 06/24/2016 10:45

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062116
Analyst1: MES Analyst2: NA
Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:
Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG573365

Internal STD: STD76627 Surrogate STD: NA Calibration STD

CCV STD: STD74784 LCS STD: STD74214 MS/MSD STD:

Comments:

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	15M18455	BAKEOUT	1	1		06/21/16 08:46
2	15M18456	WG573501-01 5PPM DFTPP STD	1	1	STD75016	06/21/16 09:05
3	15M18457	WG573501-01 5PPM DFTPP STD	1	1	STD75016	06/21/16 09:34
4	15M18458	WG573501-02 5PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 09:52
5	15M18459	WG573501-03 10 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 10:27
6	15M18460	WG573501-04 7.5 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 10:49
7	15M18461	WG573501-05 2.5 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 11:12
8	15M18462	WG573501-06 1 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 11:34
9	15M18463	WG573501-07 0.4 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 11:57
10	15M18464	WG573501-08 5PPM 1,4-DIOXANE ALT SRC	1	1	STD74214	06/21/16 12:20
11	15M18465	WG572806-01 BLANK 6/15	1	1		06/21/16 12:54
12	15M18466	WG572806-02 LCS 6/15	1	1		06/21/16 13:17
13	15M18467	WG572806-03 LCSDUP 6/15	1	1		06/21/16 13:40
14	15M18468	WG573008-01 BLANK 6/17	7	1	SOIL	06/21/16 14:03
15	15M18469	WG573008-02 LCS 6/17	7	1	SOIL	06/21/16 14:26
16	15M18470	WG573008-03 LCSDUP 6/17	7	1	SOIL	06/21/16 14:48
17	15M18471	L16060730-01	1	1		06/21/16 15:11
18	15M18472	L16060730-02	1	1		06/21/16 15:34
19	15M18473	L16060854-01	1	1		06/21/16 15:57
20	15M18474	L16060730-03	1	1		06/21/16 16:19
21	15M18475	L16060756-04 827-DIOX	1	1		06/21/16 16:46
22	15M18476	L16060822-01	1	1		06/21/16 17:08
23	15M18477	L16060822-03	1	1		06/21/16 17:31
24	15M18478	L16060822-04	1	1		06/21/16 17:54
25	15M18479	L16060822-05	1	1		06/21/16 18:17
26	15M18480	L16060823-01	1	1		06/21/16 18:40
27	15M18481	L16060823-03	1	1		06/21/16 19:02
28	15M18482	L16060823-04	1	1		06/21/16 19:25
29	15M18483	L16060809-03 827-DIOX	1	1		06/21/16 19:48
30	15M18484	L16060854-01 2X	1	2		06/21/16 20:10
31	15M18485	BAKEOUT	1	1		06/21/16 20:33
32	15M18486	BAKEOUT	1	1		06/21/16 20:55
33	15M18487	BAKEOUT	1	1		06/21/16 21:18

Page: 1 Approved: 22-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062116

Analyst1: MES Analyst2: NA

Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:

Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG573365

Internal STD: STD76627 Surrogate STD: NA

CCV STD: STD74784 LCS STD: STD74214

Comments

Seq.	Rerun	Dil.	Reason	Analytes
2				
	WG573501-01 5PPM DFTPP STD rerun DFTPP, pentachlorophenol tailing >2.			
14				
	WG573008-01 BLANK 6/17 DNR-not needed.			
15				
	WG573008-02 LCS DNR-not needed.			
16				
	WG573008-03 LCSDUP DNR-not needed.			
19	X	2	Over Calibration Range	
	L16060854-01			

Page: 2 Approved: 22-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062316
Analyst1: MES Analyst2: NA
Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:
Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG573853

Internal STD: STD76627 Surrogate STD: NA Calibration STD
CCV STD: STD74784 LCS STD: MS/MSD STD:

Comments:

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	15M18504	BAKEOUT	1	1		06/23/16 10:23
2	15M18505	WG573879-01 5PPM DFTPP STD	1	1	STD75016	06/23/16 10:43
3	15M18506	WG573879-01 5PPM DFTPP STD	1	1	STD75016	06/23/16 10:59
4	15M18507	WG573879-01 5PPM DFTPP STD	1	1	STD75016	06/23/16 12:04
5	15M18508	WG573879-01 5PPM DFTPP STD	1	1	STD75016	06/23/16 13:20
6	15M18509	WG573879-01 5PPM DFTPP STD	1	1	STD75016	06/23/16 13:42
7	15M18510	WG573879-02 5PPM 1,4-DIOXANE STD	1	1	STD74784	06/23/16 13:59
8	15M18511	WG573299-01 BLANK 6/21	1	1		06/23/16 14:22
9	15M18512	WG573299-02 LCS 6/21	1	1		06/23/16 14:45
10	15M18513	WG573299-03 LCSDUP 6/21	1	1		06/23/16 15:07
11	15M18514	L16060927-01 827-DIOX	1	1		06/23/16 15:30
12	15M18515	L16060927-02 827-DIOX	1	1		06/23/16 15:53
13	15M18516	L16060927-03 827-DIOX	1	1		06/23/16 16:15
14	15M18517	L16060927-04 827-DIOX	1	1		06/23/16 16:38
15	15M18518	L16060930-03 827-DIOX	1	1		06/23/16 17:00
16	15M18519	L16060978-03 827-DIOX	1	1		06/23/16 17:23
17	15M18520	L16060978-04 827-DIOX	1	1		06/23/16 17:46
18	15M18521	L16061026-04 827-DIOX	1	1		06/23/16 18:08
19	15M18522	L16061026-05 827-DIOX	1	1		06/23/16 18:31
20	15M18523	L16061026-07 827-DIOX	1	1		06/23/16 18:54
21	15M18524	L16061055-04 827-DIOX	1	1		06/23/16 19:16
22	15M18525	L16061055-07 827-DIOX	1	1		06/23/16 19:39
23	15M18526	BAKEOUT	1	1		06/23/16 20:02
24	15M18527	BAKEOUT	1	1		06/23/16 20:24
25	15M18528	BAKEOUT	1	1		06/23/16 20:47

Comments

Seq.	Rerun	Dil.	Reason	Analytes
2				
	Rerun DFTPP, degradation >20%.			
3				
	Rerun DFTPP, degradation >20%. Changing the liner.			
4				

Page: 1 Approved: 24-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062316

Analyst1: MES Analyst2: NA

Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:

Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG573853

Internal STD: STD76627 Surrogate STD: NA

CCV STD: STD74784 LCS STD:

Comments

Seq.	Rerun	Dil.	Reason	Analytes
	Rerun DFTPP, degradation still >20%. Changing the gold seal.			
5				
	Rerun DFTPP, degradation still >20%. Clipping the column.			
12	X	10	Surrogate standard failure	
	L16060927-02 high surrogate due to SMI; reanalyzing at a dilution to confirm.			
16	X	10	Over Calibration Range	
	L16060978-03			
17	X	20	Over Calibration Range	
	L16060978-04			
19	X	20	Over Calibration Range	
	L16061026-05 high surrogate due to SMI.			
21	X	10	Over Calibration Range	
	L16061055-04 827-DIOX one high surrogate due to SMI.			
22	X	100	Over Calibration Range	
	L16061055-07 827-DIOX one low surrogate due to obvious SMI.			

Page: 2 Approved: 24-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062516

Analyst1: MES Analyst2: NA

Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:

Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG574187,WG573853,WG574186

Internal STD: STD76627 Surrogate STD: NA Calibration STD

CCV STD: STD74784 LCS STD: MS/MSD STD:

Comments: The following may need re-extracted due to low LCS/LCSDUP recoveries:L16060637-02, 756-01 to -03,809-01,02,930-01,-02,1213,-01 to -06

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	15M18543	BAKEOUT	1	1		06/25/16 09:07
2	15M18544	WG574029-01 5PPM DFTPP STD	1	1	STD75016	06/25/16 09:27
3	15M18545	WG574029-02 5PPM 1,4-DIOXANE STD	1	1	STD74784	06/25/16 09:44
4	15M18546	WG573815-01 BLANK 6/23	7	1	SOIL	06/25/16 10:08
5	15M18547	WG573815-02 LCS 6/23	7	1	SOIL	06/25/16 10:30
6	15M18548	WG573815-03 LCSDUP 6/23	7	1	SOIL	06/25/16 10:53
7	15M18549	WG573975-01 BLANK 6/24	7	1	SOIL	06/25/16 11:16
8	15M18550	WG573975-02 LCS 6/24	7	1	SOIL	06/25/16 11:38
9	15M18551	WG573975-03 LCSDUP 6/24	7	1	SOIL	06/25/16 12:01
10	15M18552	L16061026-07 RR 827-DIOX	1	10		06/25/16 12:24
11	15M18553	L16060927-02 10X 827-DIOX	1	10		06/25/16 12:46
12	15M18554	L16060978-03 10X 827-DIOX	1	10		06/25/16 13:09
13	15M18555	L16061055-04 10X 827-DIOX	1	10		06/25/16 13:32
14	15M18556	L16060978-04 20X 827-DIOX	1	20		06/25/16 13:54
15	15M18557	L16061026-05 20X 827-DIOX	1	20		06/25/16 14:17
16	15M18558	L16061055-07 100X 827-DIOX	1	100		06/25/16 14:40
17	15M18559	L16060637-02 RE 827-DIOX	7	1	SOIL	06/25/16 15:02
18	15M18560	L16060756-01 RE 827-DIOX	7	1	SOIL	06/25/16 15:25
19	15M18561	L16060756-02 RE 827-DIOX	7	1	SOIL	06/25/16 15:47
20	15M18562	L16060756-03 RE 827-DIOX	7	1	SOIL	06/25/16 16:10
21	15M18563	L16060809-01 RE 827-DIOX	7	1	SOIL	06/25/16 16:33
22	15M18564	L16060809-02 RE 827-DIOX	7	1	SOIL	06/25/16 16:55
23	15M18565	L16060930-01 RE 827-DIOX	7	1	SOIL	06/25/16 17:18
24	15M18566	L16060930-02 RE 827-DIOX	7	1	SOIL	06/25/16 17:40
25	15M18567	L16060978-02 RE 2X 827-DIOX	7	2	SOIL	06/25/16 18:03
26	15M18568	L16061055-03 RE 827-DIOX	7	1	SOIL	06/25/16 18:26
27	15M18569	L16061055-05 RE 827-DIOX	7	1	SOIL	06/25/16 18:48
28	15M18570	L16061055-06 RE 827-DIOX	7	1	SOIL	06/25/16 19:11
29	15M18571	L16061213-01 827-DIOX	7	1	SOIL	06/25/16 19:34
30	15M18572	L16061213-02 827-DIOX	7	1	SOIL	06/25/16 19:56
31	15M18573	L16061213-03 827-DIOX	7	1	SOIL	06/25/16 20:19
32	15M18574	L16061213-04 827-DIOX	7	1	SOIL	06/25/16 20:42
33	15M18575	L16061213-05 5X 827-DIOX	7	5	SOIL	06/25/16 21:04

Page: 1 Approved: 27-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062516
Analyst1: MES Analyst2: NA
Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:
Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG574187,WG573853,WG574186

Internal STD: STD76627 Surrogate STD: NA

CCV STD: STD74784 LCS STD:

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
34	15M18576	L16061213-06 5X 827-DIOX	7	5	SOIL	06/25/16 21:27
35	15M18577	BAKEOUT	1	1		06/25/16 21:49
36	15M18578	BAKEOUT	1	1		06/25/16 22:11
37	15M18579	BAKEOUT	1	1		06/25/16 22:34
38	15M18580	BAKEOUT	1	1		06/25/16 22:56
39	15M18581	BAKEOUT	1	1		06/25/16 23:19
40	15M18582	BAKEOUT	1	1		06/25/16 23:42
41	15M18583	BAKEOUT	1	1		06/26/16 00:04

Comments

Seq.	Rerun	Dil.	Reason	Analytes
5				
	WG573815-02 LCS 6/23 low remix and rerun.			
6				
	WG573815-03 LCSDUP 6/23 low remix and rerun.			
9				
	WG573975-03 LCSDUP 6/24 marginally low; rerun. Rerun was acceptable. Not report this run.			
10				
	L16061026-07 RR 827-DIOX DNR-confirms previous results.			
11				
	L16060927-02 10X 827-DIOX surrogate is good; indicates SMI in the straight run.			
13				
	L16061055-04 10X 827-DIOX surrogate is good; indicates SMI in the straight run.			
15	X	40	Over Calibration Range	
	L16061026-05 20X 827-DIOX DNR			
17				
	L16060637-02 RE 827-DIOX surrogate is good; does not confirm low surrogate in original extraction.			
18				
	L16060756-01 RE 827-DIOX low surrogate recovery; does not confirm original extraction. Needs re-extracted.			
33	X	1	Analyzed too dilute	
	L16061213-05 5X 827-DIOX DNR			
34	X	1	Analyzed too dilute	
	L16061213-06 5X 827-DIOX DNR			

Page: 2 Approved: 27-JUN-16

Microbac Laboratories Inc.

Data Checklist

Date: 21-JUN-2016
Analyst: MES
Analyst: NA
Method: 827-DIOX
Instrument: HPMS15
Curve Workgroup: NA
Runlog ID: 75850
Analytical Workgroups: L16060730, 060756, 060809, 060822, 060823, 060854

ANALYTICAL	
System Performance Check	X
DFTPP (MS)	X
Endrin/DDT breakdown (8081/MS)	X
Pentachlorophenol/benzidine tailing (MS)	X
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	X
Average RF	X
Linear regression or higher order curve	NA
Alternate source standard (ICV) % Difference	X
Continuing Calibration (CCV)	NA
% D/% Drift	NA
Minimum response factors (MS)	NA
Continuing calibration blank (CCB) (IC)	NA
Special standards	NA
Blanks	X
TCL hits	X
Surrogate recoveries	X
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	X
MS/MSD/Sample duplicates	NA
Recoveries	NA
%RPD	NA
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	X
Surrogate recoveries	X
Internal standard areas (MS)	X
Library searches (MS)	NA
Calculations & correct factors	X
Compounds above calibration range	X
Reruns	X
Manual integrations	NA
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	NA
Check for completeness	X
Primary Reviewer	MES
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	ECL

Primary Reviewer: 22-JUN-2016

Secondary Reviewer: 22-JUN-2016

CHECKLIST1 - Modified 03/05/2008

Generated: JUN-22-2016 16:00:51

Microbac Laboratories Inc.

Data Checklist

Date: 23-JUN-2016
Analyst: MES
Analyst: NA
Method: 827-DIOX
Instrument: HPMS15
Curve Workgroup: NA
Runlog ID: 75899
Analytical Workgroups: L16060927,060930,060978,-061026,061055

ANALYTICAL	
System Performance Check	X
DFTPP (MS)	X
Endrin/DDT breakdown (8081/MS)	X
Pentachlorophenol/benzidine tailing (MS)	X
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	NA
Average RF	NA
Linear regression or higher order curve	NA
Alternate source standard (ICV) % Difference	NA
Continuing Calibration (CCV)	X
% D/% Drift	X
Minimum response factors (MS)	X
Continuing calibration blank (CCB) (IC)	NA
Special standards	NA
Blanks	X
TCL hits	X
Surrogate recoveries	X
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	X
MS/MSD/Sample duplicates	NA
Recoveries	NA
%RPD	NA
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	X
Surrogate recoveries	X
Internal standard areas (MS)	X
Library searches (MS)	NA
Calculations & correct factors	X
Compounds above calibration range	X
Reruns	X
Manual integrations	X
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	NA
Check for completeness	X
Primary Reviewer	MES
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	WTD

Primary Reviewer: Secondary Reviewer: 24-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-24-2016 14:48:11

Microbac Laboratories Inc.

Data Checklist

Date: 25-JUN-2016
Analyst: MES
Analyst: NA
Method: 827-DIOX
Instrument: HPMS15
Curve Workgroup: NA
Runlog ID: 75935
Analytical Workgroups: 060927, 06-978, 06-1055, 06-637, 06-756, 06-809, 06-930, 06-1213

ANALYTICAL	
System Performance Check	X
DFTPP (MS)	X
Endrin/DDT breakdown (8081/MS)	X
Pentachlorophenol/benzidine tailing (MS)	X
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	NA
Average RF	NA
Linear regression or higher order curve	NA
Alternate source standard (ICV) % Difference	NA
Continuing Calibration (CCV)	X
% D/% Drift	X
Minimum response factors (MS)	X
Continuing calibration blank (CCB) (IC)	NA
Special standards	NA
Blanks	X
TCL hits	X
Surrogate recoveries	X
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	X
MS/MSD/Sample duplicates	NA
Recoveries	NA
%RPD	NA
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	X
Surrogate recoveries	X
Internal standard areas (MS)	X
Library searches (MS)	NA
Calculations & correct factors	X
Compounds above calibration range	X
Reruns	X
Manual integrations	X
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	NA
Check for completeness	X
Primary Reviewer	MES
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	ECL

Primary Reviewer: 27-JUN-2016

Secondary Reviewer: 27-JUN-2016

CHECKLIST1 - Modified 03/05/2008

Generated: JUN-27-2016 14:41:23

Microbac Laboratories Inc.

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 8270D

AAB#: WG573853

Login Number: L16060927

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18CPTMW23SW-061516	01	06/15/16					06/21/2016	6.4	7		06/23/16	1.9	40	
18CPTMW23-061516	02	06/15/16					06/21/2016	6.3	7		06/23/16	2	40	
18CPTMW23-061516	02	06/15/16					06/21/2016	6.3	7		06/25/16	3.8	40	
18CPTMW19SW-061516	03	06/15/16					06/21/2016	6.3	7		06/23/16	2	40	
18CPTMW19-061516	04	06/15/16					06/21/2016	6.2	7		06/23/16	2	40	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4822545
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.
SURROGATE STANDARDS

Login Number: L16060927
Instrument Id: HPMS15
Workgroup (AAB#): WG573853

Method: 827-DIOXANE
CAL ID: HPMS15-21-JUN-16
Matrix: Water

Sample Number	Dilution	Tag	1
L16060927-01	1.00	01	32.4
L16060927-02	1.00	01	133
L16060927-02	10.0	DL01	109
L16060927-03	1.00	01	37.5
L16060927-04	1.00	01	40.5
WG573299-01	1.00	01	28.7
WG573299-02	1.00	01	37.8
WG573299-03	1.00	01	46.8

Surrogates	Surrogate Limits
1 - 1,4-Dioxane-d8	20 - 129

Underline = Result out of surrogate limits

DL = surrogate diluted out
ND = surrogate not detected

SURROGATES - Modified 03/06/2008
PDF File ID: 4822553
Report generated: 06/28/2016 09:16

METHOD BLANK SUMMARY

Login Number: L16060927
Blank File ID: 15M18511
Prep Date: 06/21/16 17:01
Analyzed Date: 06/23/16 14:22
Analyst: MES

Work Group: WG573853
Blank Sample ID: WG573299-01
Instrument ID: HPMS15
Method: 8270D

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573299-02	15M18512	06/23/16 14:45	01
LCS2	WG573299-03	15M18513	06/23/16 15:07	01
18CPTMW23SW-061516	L16060927-01	15M18514	06/23/16 15:30	01
18CPTMW23-061516	L16060927-02	15M18515	06/23/16 15:53	01
18CPTMW19SW-061516	L16060927-03	15M18516	06/23/16 16:15	01
18CPTMW19-061516	L16060927-04	15M18517	06/23/16 16:38	01
18CPTMW23-061516	L16060927-02	15M18553	06/25/16 12:46	DL01

Report Name: BLANK_SUMMARY
PDF File ID: 4823844
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060927 Prep Date: 06/21/16 17:01 Sample ID: WG573299-01
Instrument ID: HPMS15 Run Date: 06/23/16 14:22 Prep Method: 3520C
File ID: 15M18511 Analyst: MES Method: 8270D
Workgroup (AAB#): WG573853 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS15-21-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
1,4-Dioxane	0.500	2.00	0.500	1	U

Surrogates	% Recovery	Surrogate Limits	Qualifier
1,4-Dioxane-d8	28.7	20 - 129	PASS

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name:BLANK
PDF ID: 4822546
28-JUN-2016 09:16

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060927 Analyst: MES Prep Method: 3520C
Instrument ID: HPMS15 Matrix: Water Method: 8270D
Workgroup (AAB#): WG573853 Units: ug/L
QC Key: DOD4 Lot #: STD75281

Sample ID: WG573299-02 LCS File ID: 15M18512 Run Date: 06/23/2016 14:45
Sample ID: WG573299-03 LCS2 File ID: 15M18513 Run Date: 06/23/2016 15:07

Analytes	LCS			LCS2			%RPD	%Rec Limits	RPD Lmt	Q
	Known	Found	% REC	Known	Found	% REC				
1,4-Dioxane	5.00	1.88	37.6	5.00	2.32	46.4	20.8	30 - 104	30	

Surogates	LCS	LCS2	Surrogate Limits	Qualifier
	% Recovery	% Recovery		
1,4-Dioxane-d8	37.8	46.8	20 - 129	PASS

* EXCEEDS %REC LIMIT

\# EXCEEDS RPD LIMIT

LCS_LCS2 - Modified 03/06/2008
PDF File ID: 4822547
Report generated: 06/28/2016 09:16

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

DFTPP

Login Number: L16060927
Instrument: HPMS15
Analyst: MES
Workgroup: WG573501

Tune ID: WG573501-01
Run Date: 06/21/2016
Run Time: 09:34
File ID: 15M18457
Cal ID: HPMS15 - 21-JUN-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51.0	198	30.0	60.0	38.8	34488	PASS
68.0	69.0	0	2.00	1.36	500	PASS
69.0	198	0	100	41.4	36833	PASS
70.0	69.0	0	2.00	0.467	172	PASS
127	198	40.0	60.0	51.9	46131	PASS
197	198	0	1.00	0	0	PASS
198	198	100	100	100	88880	PASS
199	198	5.00	9.00	6.66	5918	PASS
275	198	10.0	30.0	24.9	22090	PASS
365	198	1.00	100	2.65	2358	PASS
441	443	0.0100	100	75.7	9586	PASS
442	198	40.0	100	73.4	65261	PASS
443	442	17.0	23.0	19.4	12659	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573501-02	STD-CCV	01	06/21/2016 09:52	
WG573501-03	STD	01	06/21/2016 10:27	
WG573501-04	STD	01	06/21/2016 10:49	
WG573501-05	STD	01	06/21/2016 11:12	
WG573501-06	STD	01	06/21/2016 11:34	
WG573501-07	STD	01	06/21/2016 11:57	
WG573501-08	SSCV	01	06/21/2016 12:20	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4822550
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

DFTPP

Login Number: L16060927
Instrument: HPMS15
Analyst: MES
Workgroup: WG573879

Tune ID: WG573879-01
Run Date: 06/23/2016
Run Time: 13:42
File ID: 15M18509
Cal ID: HPMS15 - 21-JUN-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51.0	198	30.0	60.0	37.0	43111	PASS
68.0	69.0	0	2.00	1.10	506	PASS
69.0	198	0	100	39.4	45795	PASS
70.0	69.0	0	2.00	0.157	72	PASS
127	198	40.0	60.0	51.3	59752	PASS
197	198	0	1.00	0	0	PASS
198	198	100	100	100	116364	PASS
199	198	5.00	9.00	7.09	8249	PASS
275	198	10.0	30.0	24.9	28968	PASS
365	198	1.00	100	2.82	3278	PASS
441	443	0.0100	100	79.2	13093	PASS
442	198	40.0	100	74.3	86464	PASS
443	442	17.0	23.0	19.1	16541	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573879-02	CCV	01	06/23/2016 13:59	
WG573299-01	BLANK	01	06/23/2016 14:22	
WG573299-02	LCS	01	06/23/2016 14:45	
WG573299-03	LCS2	01	06/23/2016 15:07	
L16060927-01	18CPTMW23SW-061516	01	06/23/2016 15:30	
L16060927-02	18CPTMW23-061516	01	06/23/2016 15:53	
L16060927-03	18CPTMW19SW-061516	01	06/23/2016 16:15	
L16060927-04	18CPTMW19-061516	01	06/23/2016 16:38	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4822550
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

DFTPP

Login Number: L16060927
Instrument: HPMS15
Analyst: MES
Workgroup: WG574029

Tune ID: WG574029-01
Run Date: 06/25/2016
Run Time: 09:27
File ID: 15M18544
Cal ID: HPMS15 - 21-JUN-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51.0	198	30.0	60.0	30.5	29885	PASS
68.0	69.0	0	2.00	1.68	556	PASS
69.0	198	0	100	33.9	33149	PASS
70.0	69.0	0	2.00	0.477	158	PASS
127	198	40.0	60.0	47.2	46146	PASS
197	198	0	1.00	0	0	PASS
198	198	100	100	100	97832	PASS
199	198	5.00	9.00	6.64	6499	PASS
275	198	10.0	30.0	26.5	25877	PASS
365	198	1.00	100	2.93	2871	PASS
441	443	0.0100	100	78.3	12355	PASS
442	198	40.0	100	85.7	83872	PASS
443	442	17.0	23.0	18.8	15788	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG574029-02	CCV	01	06/25/2016 09:44	
L16060927-02	18CPTMW23-061516	DL01	06/25/2016 12:46	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4822550
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.

INITIAL CALIBRATION SUMMARY

Login Number: L16060927
Analytical Method: 8270D
ICAL Workgroup: WG573501

Instrument ID: HPMS15
Initial Calibration Date: 21-JUN-16 11:57
Column ID: F

Analyte		AVG RF	% RSD	LINEAR (R)	QUAD (R^2)
1,4-Dioxane		0.3060	2.56		

R = Correlation coefficient; 0.995 minimum
R^2 = Coefficient of determination; 0.99 minimum

INT_CAL - Modified 03/06/2008
PDF File ID: 4822548
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16060927
Analytical Method: 8270D

Instrument ID: HPMS15
Initial Calibration Date: 21-JUN-16 11:57
Column ID: F

Analyte	WG573501-02			WG573501-03			WG573501-04		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
1,4-Dioxane	5.00	60883.0000	0.3073	10.0	138820.000	0.3164	7.50	103182.000	0.3106

INT_CAL - Modified 03/06/2008
PDF File ID: 4822548
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16060927
Analytical Method: 8270D

Instrument ID: HPMS15
Initial Calibration Date: 21-JUN-16 11:57
Column ID: F

Analyte	WG573501-05			WG573501-06			WG573501-07		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
1,4-Dioxane	2.50	26333.0000	0.2969	1.00	13250.0000	0.2966	0.400	5650.00000	0.3081

INT_CAL - Modified 03/06/2008
PDF File ID: 4822548
Report generated 06/28/2016 09:16

Calibration Table Report
Method: DIOXANE_D8.M
Title: OVD MSS01 SIM 1,4-dioxane ICAL 062116
Last Calibration: Tue Jun 21 12:51:17 2016
Curve:WG573501
Calibration Files

Compound	10 15M18459.D	7.5 15M18460.D	5 15M18458.D	2.5 15M18461.D	1 15M18462.D	0.4 15M18463.D	Avg	%RSD
1,4-Dichlorobenzene-d4	ISTD							
1,4-Dioxane	0.316	0.311	0.307	0.297	0.297	0.308	0.306	2.558
1,4-Dioxane-d8	0.426	0.418	0.417	0.402	0.395		0.411	3.044
Nitrobenzene-d5	1.202	1.173	1.128	1.002	0.927	0.906	1.056	12.163
2-Fluorobiphenyl	2.735	2.736	2.691	2.621	2.563	2.651	2.666	2.548
p-Terphenyl-d14	2.793	2.774	2.622	2.460	2.367	2.405	2.570	7.266

Wed Jun 22 12:17:33 2016

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16060927 Run Date: 06/21/2016 Sample ID: WG573501-08
Instrument ID: HPMS15 Run Time: 12:20 Method: 8270D
File ID: 15M18464 Analyst: MES QC Key: DOD4
ICal Workgroup: WG573501 Cal ID: HPMS15 - 21-JUN-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
1,4-Dioxane		5000	5050	ug/L	0.309	0.900	20	

* Exceeds %D Limit

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4822549
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927
Run Date: 06/23/2016
Sample ID: WG573879-02
Instrument ID: HPMS15
Run Time: 13:59
Method: 8270D
File ID: 15M18510
Analyst: MES
QC Key: DOD4
Workgroup (AAB#): WG573853
Cal ID: HPMS15 - 21-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
1,4-Dioxane		5000	4530	ug/L	0.277	9.36	20	

* Exceeds %D Criteria

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

CCV - Modified 03/05/2008
PDF File ID: 4822551
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927
Run Date: 06/25/2016
Sample ID: WG574029-02
Instrument ID: HPMS15
Run Time: 09:44
Method: 8270D
File ID: 15M18545
Analyst: MES
QC Key: DOD4
Workgroup (AAB#): WG573853
Cal ID: HPMS15 - 21-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
1,4-Dioxane		5000	4170	ug/L	0.255	16.6	20	

* Exceeds %D Criteria

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

CCV - Modified 03/05/2008
PDF File ID: 4822551
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060927
Instrument ID: HPMS15
Workgroup (AAB#): WG573853

ICAL CCV Number: WG573501-02
CAL ID: HPMS15-21-JUN-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1
WG573501-02	NA	NA	39637
Upper Limit	NA	NA	79274
Lower Limit	NA	NA	19819
L16060927-01	1.00	01	52572
L16060927-02	1.00	01	52222
L16060927-02	10.0	DL01	44449
L16060927-03	1.00	01	56209
L16060927-04	1.00	01	51577
WG573299-01	1.00	01	57608
WG573299-02	1.00	01	51597
WG573299-03	1.00	01	54701

IS-1 - 1,4-Dichlorobenzene-d4

Underline = Response outside limits

INTERNAL_STD_ICAL - Modified 03/06/2008
PDF File ID: 4822552
Report generated 06/28/2016 09:16

Microbac Laboratories Inc.
INTERNAL STANDARD RETENTION TIME SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16060927
Instrument ID: HPMS15
Workgroup (AAB#): WG573853

ICAL CCV Number: WG573501-02
CAL ID: HPMS15-21-JUN-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1
WG573501-02	NA	NA	6.93
Upper Limit	NA	NA	7.43
Lower Limit	NA	NA	6.43
L16060927-01	1.00	01	6.933
L16060927-02	1.00	01	6.929
L16060927-02	10.0	DL01	6.896
L16060927-03	1.00	01	6.933
L16060927-04	1.00	01	6.933
WG573299-01	1.00	01	6.929
WG573299-02	1.00	01	6.929
WG573299-03	1.00	01	6.929

IS-1 - 1,4-Dichlorobenzene-d4

Underline = Response outside limits

INTERNAL_STD_RT_ICAL - Modified 03/06/2008
PDF File ID: 4823845
Report generated: 06/28/2016 09:16

2.2.1.3 Sample Data

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18514.D
Acq On : 23 Jun 2016 15:30
Operator : MES
Sample : L16060927-01 827-DIOX
Misc : 1,1
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jun 23 15:52:27 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	52572	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.862	96	35041	1.6198	ug/L	0.04
Spiked Amount 5.000			Recovery	=	32.40%	
4) Nitrobenzene-d5	7.431	82	2549	0.0459	ug/L	0.00
Spiked Amount 5.000			Recovery	=	1.00%	
5) 2-Fluorobiphenyl	8.919	172	664	0.0047	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.00%	
6) p-Terphenyl-d14	12.166	244	763	0.0056	ug/L	-0.02
Spiked Amount 5.000			Recovery	=	0.20%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18514.D
Acq On : 23 Jun 2016 15:30
Operator : MES
Sample : L16060927-01 827-DIOX
Misc : 1,1
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jun 23 15:52:27 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (QT Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18515.D
Acq On : 23 Jun 2016 15:53
Operator : MES
Sample : L16060927-02 827-DIOX
Misc : 1,1
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jun 24 10:29:50 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.929	152	52222	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.848	96	142971m	6.6534	ug/L	0.02
Spiked Amount 5.000			Recovery	= 133.00%		
4) Nitrobenzene-d5	7.431	82	3760	0.0682	ug/L	0.00
Spiked Amount 5.000			Recovery	= 1.40%		
5) 2-Fluorobiphenyl	8.919	172	1557	0.0112	ug/L	0.00
Spiked Amount 5.000			Recovery	= 0.20%		
6) p-Terphenyl-d14	12.166	244	949	0.0071	ug/L	-0.02
Spiked Amount 5.000			Recovery	= 0.20%		
Target Compounds						Qvalue
2) 1,4-Dioxane	3.897	58	121747	7.6196	ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18515.D
Acq On : 23 Jun 2016 15:53
Operator : MES
Sample : L16060927-02 827-DIOX
Misc : 1,1
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jun 24 10:29:50 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 7.62 ug/L
RT: 3.897 min Scan# 116
Delta R.T. 0.035 min
Lab File: 15M18515.D
Acq: 23 Jun 2016 15:53

Tgt Ion: 58 Resp: 121747

Ion	Ratio	Lower	Upper
58	100		
88	152.0	103.1	240.7



Data Path : C:\msdchem\1\data\062316\
Data File : 15M18515.D
Acq On : 23 Jun 2016 15:53
Operator : MES
Sample : L16060927-02 827-DIOX
Misc : 1,1
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jun 23 17:16:01 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



TIC: 15M18515.D\data.ms

(3) 1,4-Dioxane-d8 (S)

3.771min (-0.056) 21.96 ug/L

response 471929

Ion	Exp%	Act%
96.10	100.00	100.00
64.10	74.80	0.00#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18515.D
Acq On : 23 Jun 2016 15:53
Operator : MES
Sample : L16060927-02 827-DIOX
Misc : 1,1
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jun 23 17:16:01 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration







TIC: 15M18515.D\data.ms

(3) 1,4-Dioxane-d8 (S)

3.848min (+ 0.021) 6.65 ug/L mint

response 142971

Ion	Exp%	Act%
96.10	100.00	100.00
64.10	74.80	0.00#
0.00	0.00	0.00
0.00	0.00	0.00

Analyst: 06/24/2016 12:00 Supervisor: 06/24/2016 14:34

#2 - Data system splits the peak incorrectly or integrates a false peak as a rider peak

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062516\
Data File : 15M18553.D
Acq On : 25 Jun 2016 12:46
Operator : MES
Sample : L16060927-02 10X 827-DIOX
Misc : 1,10
ALS Vial : 10 Sample Multiplier: 10

Quant Time: Jun 26 17:44:43 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.896	152	44449	1.0000	ug/mL	#-0.04
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.764	96	9927	5.4275	ug/L	-0.06
Spiked Amount 5.000			Recovery	=	108.60%	
4) Nitrobenzene-d5	7.404	82	289	0.0616	ug/L	-0.03
Spiked Amount 5.000			Recovery	=	1.20%	
5) 2-Fluorobiphenyl	8.890	172	148	0.0125	ug/L	-0.03
Spiked Amount 5.000			Recovery	=	0.20%	
6) p-Terphenyl-d14	12.085	244	5	0.0004	ug/L	-0.10
Spiked Amount 5.000			Recovery	=	0.00%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.799	58	10910	0.8022	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062516\
Data File : 15M18553.D
Acq On : 25 Jun 2016 12:46
Operator : MES
Sample : L16060927-02 10X 827-DIOX
Misc : 1,10
ALS Vial : 10 Sample Multiplier: 10

Quant Time: Jun 26 17:44:43 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 0.80 ug/L
RT: 3.799 min Scan# 102
Delta R.T. -0.063 min
Lab File: 15M18553.D
Acq: 25 Jun 2016 12:46

Tgt Ion: 58 Resp: 10910

Ion	Ratio	Lower	Upper
58	100		
88	149.2	103.1	240.7



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18516.D
Acq On : 23 Jun 2016 16:15
Operator : MES
Sample : L16060927-03 827-DIOX
Misc : 1,1
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 23 17:16:26 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	56209	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.862	96	43386	1.8758	ug/L	0.04
Spiked Amount 5.000			Recovery	=	37.60%	
4) Nitrobenzene-d5	7.431	82	3380	0.0569	ug/L	0.00
Spiked Amount 5.000			Recovery	=	1.20%	
5) 2-Fluorobiphenyl	8.919	172	1644	0.0110	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.20%	
6) p-Terphenyl-d14	12.166	244	2195	0.0152	ug/L	-0.02
Spiked Amount 5.000			Recovery	=	0.40%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18516.D
Acq On : 23 Jun 2016 16:15
Operator : MES
Sample : L16060927-03 827-DIOX
Misc : 1,1
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 23 17:16:26 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18517.D
Acq On : 23 Jun 2016 16:38
Operator : MES
Sample : L16060927-04 827-DIOX
Misc : 1,1
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Jun 23 17:16:29 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	51577	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.862	96	42956	2.0240	ug/L	0.04
Spiked Amount 5.000			Recovery	=	40.40%	
4) Nitrobenzene-d5	7.216	82	1787	0.0328	ug/L	-0.21
Spiked Amount 5.000			Recovery	=	0.60%	
5) 2-Fluorobiphenyl	8.919	172	341	0.0025	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.00%	
6) p-Terphenyl-d14	12.166	244	2435	0.0184	ug/L	-0.02
Spiked Amount 5.000			Recovery	=	0.40%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18517.D
Acq On : 23 Jun 2016 16:38
Operator : MES
Sample : L16060927-04 827-DIOX
Misc : 1,1
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Jun 23 17:16:29 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



2.2.1.4 Standards Data

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18458.D
Acq On : 21 Jun 2016 9:52
Operator : MES
Sample : WG573501-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 22 08:04:39 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	6.933	152	39637	1.0000 ug/mL	# 0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.827	96	82573	5.0627 ug/L	0.00
Spiked Amount 5.000			Recovery	= 101.20%	
4) Nitrobenzene-d5	7.431	82	223381	5.3353 ug/L	0.00
Spiked Amount 5.000			Recovery	= 106.80%	
5) 2-Fluorobiphenyl	8.922	172	533101	5.0444 ug/L	0.00
Spiked Amount 5.000			Recovery	= 100.80%	
6) p-Terphenyl-d14	12.184	244	519544	5.1003 ug/L	0.00
Spiked Amount 5.000			Recovery	= 102.00%	
Target Compounds					Qvalue
2) 1,4-Dioxane	3.862	58	60883	5.0202 ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18458.D
Acq On : 21 Jun 2016 9:52
Operator : MES
Sample : WG573501-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 22 08:04:39 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18459.D
Acq On : 21 Jun 2016 10:27
Operator : MES
Sample : WG573501-03 10 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 22 08:04:42 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	43874	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.834	96	186767	10.3452	ug/L	0.00
Spiked Amount 5.000			Recovery = 207.00%			
4) Nitrobenzene-d5	7.431	82	527549	11.3834	ug/L	0.00
Spiked Amount 5.000			Recovery = 227.60%			
5) 2-Fluorobiphenyl	8.926	172	1199799	10.2565	ug/L	0.00
Spiked Amount 5.000			Recovery = 205.20%			
6) p-Terphenyl-d14	12.240	244	1225328	10.8672	ug/L	0.06
Spiked Amount 5.000			Recovery = 217.40%			
Target Compounds						Qvalue
2) 1,4-Dioxane	3.869	58	138820	10.3412	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18459.D
Acq On : 21 Jun 2016 10:27
Operator : MES
Sample : WG573501-03 10 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 22 08:04:42 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18460.D
Acq On : 21 Jun 2016 10:49
Operator : MES
Sample : WG573501-04 7.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 22 08:04:45 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	44299	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.834	96	138833	7.6163	ug/L	0.00
Spiked Amount 5.000			Recovery	= 152.40%		
4) Nitrobenzene-d5	7.431	82	389810	8.3306	ug/L	0.00
Spiked Amount 5.000			Recovery	= 166.60%		
5) 2-Fluorobiphenyl	8.923	172	909050	7.6965	ug/L	0.00
Spiked Amount 5.000			Recovery	= 154.00%		
6) p-Terphenyl-d14	12.203	244	921380	8.0931	ug/L	0.02
Spiked Amount 5.000			Recovery	= 161.80%		
Target Compounds						Qvalue
2) 1,4-Dioxane	3.876	58	103182	7.6127	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18460.D
Acq On : 21 Jun 2016 10:49
Operator : MES
Sample : WG573501-04 7.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 22 08:04:45 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18461.D
Acq On : 21 Jun 2016 11:12
Operator : MES
Sample : WG573501-05 2.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jun 22 08:04:48 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	35478	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.827	96	35630	2.4406	ug/L	0.00
Spiked Amount 5.000			Recovery = 48.80%			
4) Nitrobenzene-d5	7.431	82	88869	2.3714	ug/L	0.00
Spiked Amount 5.000			Recovery = 47.40%			
5) 2-Fluorobiphenyl	8.923	172	232500	2.4579	ug/L	0.00
Spiked Amount 5.000			Recovery = 49.20%			
6) p-Terphenyl-d14	12.199	244	218191	2.3930	ug/L	0.02
Spiked Amount 5.000			Recovery = 47.80%			
Target Compounds						Qvalue
2) 1,4-Dioxane	3.862	58	26333	2.4259	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18461.D
Acq On : 21 Jun 2016 11:12
Operator : MES
Sample : WG573501-05 2.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jun 22 08:04:48 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18462.D
Acq On : 21 Jun 2016 11:34
Operator : MES
Sample : WG573501-06 1 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jun 22 08:04:51 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	6.933	152	44672	1.0000 ug/mL	# 0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.841	96	17660	0.9607 ug/L	0.01
Spiked Amount 5.000			Recovery	= 19.20%	
4) Nitrobenzene-d5	7.431	82	41409	0.8776 ug/L	0.00
Spiked Amount 5.000			Recovery	= 17.60%	
5) 2-Fluorobiphenyl	8.922	172	114491	0.9612 ug/L	0.00
Spiked Amount 5.000			Recovery	= 19.20%	
6) p-Terphenyl-d14	12.203	244	105713	0.9208 ug/L	0.02
Spiked Amount 5.000			Recovery	= 18.40%	
Target Compounds					Qvalue
2) 1,4-Dioxane	3.876	58	13250	0.9694 ug/L	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18462.D
Acq On : 21 Jun 2016 11:34
Operator : MES
Sample : WG573501-06 1 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jun 22 08:04:51 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18463.D
Acq On : 21 Jun 2016 11:57
Operator : MES
Sample : WG573501-07 0.4 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jun 22 08:04:54 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	45840	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.841	96	7629	0.4045	ug/L	0.01
Spiked Amount 5.000			Recovery	=	8.00%	
4) Nitrobenzene-d5	7.431	82	16604	0.3429	ug/L	0.00
Spiked Amount 5.000			Recovery	=	6.80%	
5) 2-Fluorobiphenyl	8.923	172	48617	0.3978	ug/L	0.00
Spiked Amount 5.000			Recovery	=	8.00%	
6) p-Terphenyl-d14	12.210	244	44097	0.3743	ug/L	0.03
Spiked Amount 5.000			Recovery	=	7.40%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.876	58	5650	0.4028	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18463.D
Acq On : 21 Jun 2016 11:57
Operator : MES
Sample : WG573501-07 0.4 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jun 22 08:04:54 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18464.D
Acq On : 21 Jun 2016 12:20
Operator : MES
Sample : WG573501-08 5PPM 1,4-DIOXANE ALT SRC
Misc : 1,1 STD74214
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 22 08:04:57 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	41944	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.792	96	22	0.0013	ug/L	-0.04
Spiked Amount 5.000			Recovery	=	0.00%	
4) Nitrobenzene-d5	7.436	82	56	0.0013	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.00%	
5) 2-Fluorobiphenyl	8.926	172	23	0.0002	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.00%	
6) p-Terphenyl-d14	12.214	244	128	0.0012	ug/L	0.03
Spiked Amount 5.000			Recovery	=	0.00%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.869	58	64750	5.0454	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18464.D
Acq On : 21 Jun 2016 12:20
Operator : MES
Sample : WG573501-08 5PPM 1,4-DIOXANE ALT SRC
Misc : 1,1 STD74214
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 22 08:04:57 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18464.D
Acq On : 21 Jun 2016 12:20
Operator : MES
Sample : WG573501-08 5PPM 1,4-DIOXANE ALT SRC
Misc : 1,1 STD74214
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 22 08:04:57 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	5.000	5.045	-0.9	106	0.00
3 S	1,4-Dioxane-d8	5.000	0.001	100.0#	0	-0.04
4 S	Nitrobenzene-d5	5.000	0.001	100.0#	0	0.00
5 S	2-Fluorobiphenyl	5.000	0.000	100.0#	0	0.00
6 S	p-Terphenyl-d14	5.000	0.001	100.0#	0	0.03

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18510.D
Acq On : 23 Jun 2016 13:59
Operator : MES
Sample : WG573879-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 23 14:15:55 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	6.929	152	50565	1.0000 ug/mL	# 0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.848	96	96203	4.6237 ug/L	0.02
Spiked Amount 5.000			Recovery	= 92.40%	
4) Nitrobenzene-d5	7.426	82	280398	5.2498 ug/L	0.00
Spiked Amount 5.000			Recovery	= 105.00%	
5) 2-Fluorobiphenyl	8.919	172	664952	4.9322 ug/L	0.00
Spiked Amount 5.000			Recovery	= 98.60%	
6) p-Terphenyl-d14	12.155	244	665166	5.1186 ug/L	-0.03
Spiked Amount 5.000			Recovery	= 102.40%	
Target Compounds					Qvalue
2) 1,4-Dioxane	3.883	58	70119	4.5322 ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18510.D
Acq On : 23 Jun 2016 13:59
Operator : MES
Sample : WG573879-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 23 14:15:55 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18510.D
Acq On : 23 Jun 2016 13:59
Operator : MES
Sample : WG573879-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 23 14:15:55 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.306	0.277	9.5	115	0.02
3 S	1,4-Dioxane-d8	0.411	0.381	7.3	117	0.02
4 S	Nitrobenzene-d5	1.056	1.109	-5.0	126	0.00
5 S	2-Fluorobiphenyl	2.666	2.630	1.4	125	0.00
6 S	p-Terphenyl-d14	2.570	2.631	-2.4	128	-0.03

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18510.D
Acq On : 23 Jun 2016 13:59
Operator : MES
Sample : WG573879-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 23 14:15:55 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	5.000	4.532	9.4	115	0.02
3 S	1,4-Dioxane-d8	5.000	4.624	7.5	117	0.02
4 S	Nitrobenzene-d5	5.000	5.250	-5.0	126	0.00
5 S	2-Fluorobiphenyl	5.000	4.932	1.4	125	0.00
6 S	p-Terphenyl-d14	5.000	5.119	-2.4	128	-0.03

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062516\
Data File : 15M18545.D
Acq On : 25 Jun 2016 9:44
Operator : MES
Sample : WG574029-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 25 10:02:11 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	6.900	152	49570	1.0000 ug/mL	#-0.03
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.820	96	86328	4.2323 ug/L	0.00
Spiked Amount 5.000			Recovery	= 84.60%	
4) Nitrobenzene-d5	7.399	82	272067	5.1960 ug/L	-0.03
Spiked Amount 5.000			Recovery	= 104.00%	
5) 2-Fluorobiphenyl	8.889	172	614471	4.6492 ug/L	-0.03
Spiked Amount 5.000			Recovery	= 93.00%	
6) p-Terphenyl-d14	12.133	244	671949	5.2746 ug/L	-0.05
Spiked Amount 5.000			Recovery	= 105.40%	
Target Compounds					Qvalue
2) 1,4-Dioxane	3.855	58	63262	4.1711 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062516\
Data File : 15M18545.D
Acq On : 25 Jun 2016 9:44
Operator : MES
Sample : WG574029-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 25 10:02:11 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062516\
Data File : 15M18545.D
Acq On : 25 Jun 2016 9:44
Operator : MES
Sample : WG574029-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 25 10:02:11 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	-0.03
2	1,4-Dioxane	0.306	0.255	16.7	104	0.00
3 S	1,4-Dioxane-d8	0.411	0.348	15.3	105	0.00
4 S	Nitrobenzene-d5	1.056	1.098	-4.0	122	-0.03
5 S	2-Fluorobiphenyl	2.666	2.479	7.0	115	-0.03
6 S	p-Terphenyl-d14	2.570	2.711	-5.5	129	-0.05

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062516\
Data File : 15M18545.D
Acq On : 25 Jun 2016 9:44
Operator : MES
Sample : WG574029-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 25 10:02:11 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	-0.03
2	1,4-Dioxane	5.000	4.171	16.6	104	0.00
3 S	1,4-Dioxane-d8	5.000	4.232	15.4	105	0.00
4 S	Nitrobenzene-d5	5.000	5.196	-3.9	122	-0.03
5 S	2-Fluorobiphenyl	5.000	4.649	7.0	115	-0.03
6 S	p-Terphenyl-d14	5.000	5.275	-5.5	129	-0.05

(#) = Out of Range SPCC's out = 0 CCC's out = 0

2.2.1.5 Raw QC Data

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\msdchem\1\methods\DFTPPQ.M
Title : DFTPP
Last Update : Mon Apr 27 13:16:06 2015





AutoFind: Scans 234, 235, 236; Background Corrected with Scan 228

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	38.8	34488	PASS
68	69	0.00	2	1.4	500	PASS
69	198	0.00	100	41.4	36833	PASS
70	69	0.00	2	0.5	172	PASS
127	198	40	60	51.9	46131	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	88880	PASS
199	198	5	9	6.7	5918	PASS
275	198	10	30	24.9	22090	PASS
365	198	1	100	2.7	2358	PASS
441	443	0.01	100	75.7	9586	PASS
442	198	40	100	73.4	65261	PASS
443	442	17	23	19.4	12659	PASS

Quantitation Report (Qedit)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 21 09:47:12 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18457.D\data.ms

(1) Pentachlorophenol

7.054min (+ 0.105) 0.00 ug/ml

response 30738

Ion	Exp%	Act%
265.80	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 21 09:47:12 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





(2) Benzidine

8.502min (+ 0.008) 0.00 ug/ml

response 532770

Ion	Exp%	Act%
184.10	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 21 09:47:12 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18457.D\data.ms

(3) DDT

9.486min (-0.108) 0.00 ug/ml

response 180187

Ion	Exp%	Act%
235.00	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18509.D
Acq On : 23 Jun 2016 13:42
Operator : MES
Sample : WG573879-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\msdchem\1\methods\DFTPPQ.M
Title : DFTPP
Last Update : Mon Apr 27 13:16:06 2015





AutoFind: Scans 233, 234, 235; Background Corrected with Scan 227

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	37.0	43111	PASS
68	69	0.00	2	1.1	506	PASS
69	198	0.00	100	39.4	45795	PASS
70	69	0.00	2	0.2	72	PASS
127	198	40	60	51.3	59752	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	116364	PASS
199	198	5	9	7.1	8249	PASS
275	198	10	30	24.9	28968	PASS
365	198	1	100	2.8	3278	PASS
441	443	0.01	100	79.2	13093	PASS
442	198	40	100	74.3	86464	PASS
443	442	17	23	19.1	16541	PASS

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18509.D
Acq On : 23 Jun 2016 13:42
Operator : MES
Sample : WG573879-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 23 13:53:21 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18509.D\data.ms

(1) Pentachlorophenol

7.048min (+ 0.099) 0.00 ug/ml

response 32081

Ion	Exp%	Act%
265.80	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18509.D
Acq On : 23 Jun 2016 13:42
Operator : MES
Sample : WG573879-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 23 13:53:21 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18509.D\data.ms

(2) Benzidine

8.496min (+ 0.002) 0.00 ug/ml

response 452358

Ion	Exp%	Act%
184.10	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18509.D
Acq On : 23 Jun 2016 13:42
Operator : MES
Sample : WG573879-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 23 13:53:21 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18509.D\data.ms

(3) DDT

9.594min (-9.594) 0.00 ug/ml

response 0

Ion	Exp%	Act%
235.00	100.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062516\
Data File : 15M18544.D
Acq On : 25 Jun 2016 9:27
Operator : MES
Sample : WG574029-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\msdchem\1\methods\DFTPPQ.M
Title : DFTPP
Last Update : Mon Apr 27 13:16:06 2015





AutoFind: Scans 227, 228, 229; Background Corrected with Scan 221

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	30.5	29885	PASS
68	69	0.00	2	1.7	556	PASS
69	198	0.00	100	33.9	33149	PASS
70	69	0.00	2	0.5	158	PASS
127	198	40	60	47.2	46146	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	97832	PASS
199	198	5	9	6.6	6499	PASS
275	198	10	30	26.5	25877	PASS
365	198	1	100	2.9	2871	PASS
441	443	0.01	100	78.3	12355	PASS
442	198	40	100	85.7	83872	PASS
443	442	17	23	18.8	15788	PASS

Data Path : C:\msdchem\1\data\062516\
Data File : 15M18544.D
Acq On : 25 Jun 2016 9:27
Operator : MES
Sample : WG574029-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 25 09:38:05 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18544.D\data.ms

(1) Pentachlorophenol

7.014min (+ 0.065) 0.00 ug/ml

response 23506

Ion	Exp%	Act%
265.80	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062516\
Data File : 15M18544.D
Acq On : 25 Jun 2016 9:27
Operator : MES
Sample : WG574029-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 25 09:38:05 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18544.D\data.ms

(2) Benzidine

8.456min (-0.038) 0.00 ug/ml

response 410023

Ion	Exp%	Act%
184.10	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)

Data Path : C:\msdchem\1\data\062516\
Data File : 15M18544.D
Acq On : 25 Jun 2016 9:27
Operator : MES
Sample : WG574029-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 25 09:38:05 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18544.D\data.ms

(3) DDT

9.428min (-0.166) 0.00 ug/ml mint

response 292084

Ion	Exp%	Act%
235.00	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18511.D
Acq On : 23 Jun 2016 14:22
Operator : MES
Sample : WG573299-01 BLANK 6/21
Misc : 1,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 23 15:52:24 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.929	152	57608	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.862	96	33985	1.4337	ug/L	0.04
Spiked Amount 5.000			Recovery = 28.60%			
4) Nitrobenzene-d5	7.426	82	1290209	21.2028	ug/L	0.00
Spiked Amount 5.000			Recovery = 424.00%			
5) 2-Fluorobiphenyl	8.919	172	2634137	17.1496	ug/L	0.00
Spiked Amount 5.000			Recovery = 343.00%			
6) p-Terphenyl-d14	12.162	244	4420616	29.8588	ug/L	-0.02
Spiked Amount 5.000			Recovery = 597.20%			

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18511.D
Acq On : 23 Jun 2016 14:22
Operator : MES
Sample : WG573299-01 BLANK 6/21
Misc : 1,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 23 15:52:24 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18512.D
Acq On : 23 Jun 2016 14:45
Operator : MES
Sample : WG573299-02 LCS 6/21
Misc : 1,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 23 15:52:38 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.929	152	51597	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.862	96	40084	1.8880	ug/L	0.04
Spiked Amount 5.000			Recovery	=	37.80%	
4) Nitrobenzene-d5	7.431	82	2991	0.0549	ug/L	0.00
Spiked Amount 5.000			Recovery	=	1.00%	
5) 2-Fluorobiphenyl	8.919	172	1228	0.0089	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.20%	
6) p-Terphenyl-d14	12.159	244	1342	0.0101	ug/L	-0.03
Spiked Amount 5.000			Recovery	=	0.20%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.897	58	29708	1.8818	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18512.D
Acq On : 23 Jun 2016 14:45
Operator : MES
Sample : WG573299-02 LCS 6/21
Misc : 1,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 23 15:52:38 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18513.D
Acq On : 23 Jun 2016 15:07
Operator : MES
Sample : WG573299-03 LCSDUP 6/21
Misc : 1,1
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jun 23 15:52:42 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.929	152	54701	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.862	96	52635	2.3384	ug/L	0.04
Spiked Amount 5.000			Recovery	=	46.80%	
4) Nitrobenzene-d5	7.431	82	5866	0.1015	ug/L	0.00
Spiked Amount 5.000			Recovery	=	2.00%	
5) 2-Fluorobiphenyl	8.919	172	4684	0.0321	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.60%	
6) p-Terphenyl-d14	12.151	244	1266	0.0090	ug/L	-0.03
Spiked Amount 5.000			Recovery	=	0.20%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.897	58	38804	2.3185	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18513.D
Acq On : 23 Jun 2016 15:07
Operator : MES
Sample : WG573299-03 LCSDUP 6/21
Misc : 1,1
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jun 23 15:52:42 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



2.2 General Chromatography Data

2.2.2 6850 LC/MS Data

2.2.2.1 Summary Data

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16060927-01	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	18CPTMW23SW-061516	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 19:02
Collect Date:	06/15/2016 08:00	**Dilution:**	1	**File ID:**	1LM.LM35501
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	0.407		0.400	0.200	0.100

Certificate of Analysis

Sample #:	L16060927-02	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	18CPTMW23-061516	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 19:21
Collect Date:	06/15/2016 09:05	**Dilution:**	1000	**File ID:**	1LM.LM35502
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	2310		400	200	100

Certificate of Analysis

Sample #:	L16060927-03	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	18CPTMW19SW-061516	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/24/2016 23:18
Collect Date:	06/15/2016 10:15	**Dilution:**	1	**File ID:**	1LM.LM35473
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	1.25		0.400	0.200	0.100

Certificate of Analysis

Microbac

Lab Report #: L16060927
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Sample #:	L16060927-04	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	18CPTMW19-061516	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/24/2016 23:37
Collect Date:	06/15/2016 11:15	**Dilution:**	1	**File ID:**	1LM.LM35474
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	5.31		0.400	0.200	0.100

2.2.2.2 QC Summary Data

Example Calculation 6850 - Perchlorate

Concentration from Linear Regression

Step 1: Retrieve Curve Data From Plot, y = mx + b

y = response ratio = response of analyte / response of internal standard (IS) = Rx/Ristd

x = amount ratio = concentration analyte/concentration internal standard (IS) = Cx / Cistd

m = slope from curve (1.45)

b = intercept from curve (-0.00242)

y = 1.45x + -0.00242

Step 2: Substitute the value for y

where y = 12600/226000 = 0.055752

Step 3: Solve for x

x = (y - b)/m = 0.0040119

Step 4: Solve for analyte concentration Cx

Cx = (Cis)(x) = (5 ug/L)(0.040119) = 0.200594 ug/L

Example Calculation - Water:

Slope from curve, m:	1.45
Intercept from curve, b:	-0.00242
Response of analyte, Rx:	12600
Response of Internal Standard , Ristd:	226000
Concentration of IS, Cistd (ug/L):	5.00
Response Ratio:	0.05575
Amount Ratio:	0.04012
Analyte Concentration, Cx (ug/L) :	0.200594

Example Calculation - Soil:

Analyte Concentration, Cx (ug/L):	0.20059
Amount of soil extracted (g):	5.00
Final volume of extract (mL):	50.00
Percent solids (Pct wt.)	100
Concentration in soil (ug/kg):	2.005938

Perchlorate Conductivity Check
(perchlorate1)

Conductivity Probe Calibration Check: 1380 /1410 µs/cm

Working MCT Level: 10,000 µs/cm

Sample	Conductivity (µs/cm)	Pretreatment or Dilution Needed
WG574101-01 MCT	9,990.	
-02 Blank	0.33	
-03 LCS	0.48	
-04 LCS2	0.51	
L16060823-01	9,120.	
-03	7,440.	
-04	1,943.	
L16060927-01	750.	
-02	652.	
-03	446.	
-04	162.7	
L16061026-01	6,910.	
-03	238.	
-04	965.	
-05	1,526.	
-07	220.	
L16061142-01	165.3	
-02	81.1	
-04	83.1	
-06	728.	
-07	74.2	
L16061144-01	118.6	
-02	118.7	
-03	110.1	

Analyst: John Richards

Date/Time: 06/27/16 / 20:00

DCN#119317

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 050316_JWR.TXT
Analyst1: JWR Analyst2: NA
Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254
Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG567013 (soils)

Internal STD: COA18071 Surrogate STD: NA Calibration STD STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 MS/MSD STD: STD75512

Comments: ICAL WG567320 : Alternate Source STD75512
Analytical Column : RPPX 5um (250x4.6mm)
K'Prime S/N RPPX250-02115

Samples L16041363(-05 and -10) were analyzed at dilutions based on their pre-run screen results.

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	1LM.LM34686	WG567320-01 CCB	1	1		05/03/16 15:06
2	1LM.LM34687	WG567320-02 STD (0.1 ug/L)	1	1	STD75510	05/03/16 15:25
3	1LM.LM34688	WG567320-03 STD (0.2 ug/L)	1	1	STD75510	05/03/16 15:43
4	1LM.LM34689	WG567320-04 STD (0.5 ug/L)	1	1	STD75510	05/03/16 16:02
5	1LM.LM34690	WG567320-05 STD (1.0 ug/L)	1	1	STD75510	05/03/16 16:21
6	1LM.LM34691	WG567320-06 STD (2.0 ug/L)	1	1	STD75510	05/03/16 16:40
7	1LM.LM34692	WG567320-07 STD (5.0 ug/L)	1	1	STD75510	05/03/16 16:59
8	1LM.LM34693	WG567320-08 STD (10 ug/L)	1	1	STD75510	05/03/16 17:18
9	1LM.LM34694	WG567320-09 SSCV (1.0 ug/L)	1	1	STD75512	05/03/16 17:37
10	1LM.LM34695	WG567321-01 CCB	1	1		05/03/16 17:56
11	1LM.LM34696	WG567321-02 CCV (1.0ug/L)	1	1	STD75510	05/03/16 18:15
12	1LM.LM34697	WG567013-07 MRL (2.0ug/kg)	7	1	STD75510	05/03/16 18:34
13	1LM.LM34698	WG567013-01 MCT (2.0ug/kg)	7	1	STD75512	05/03/16 18:53
14	1LM.LM34699	WG567013-02 BLANK	7	1		05/03/16 19:12
15	1LM.LM34700	WG567013-03 LCS (2.0ug/kg)	7	1	STD75512	05/03/16 19:31
16	1LM.LM34701	L16041363-07 RS	7	1		05/03/16 19:50
17	1LM.LM34702	L16041363-08 MS	7	1	STD75512	05/03/16 20:09
18	1LM.LM34703	L16041363-09 MSD	7	1	STD75512	05/03/16 20:28
19	1LM.LM34704	L16041363-01	7	1		05/03/16 20:46
20	1LM.LM34705	L16041363-02	7	1		05/03/16 21:05
21	1LM.LM34706	L16041363-03	7	1		05/03/16 21:24
22	1LM.LM34707	L16041363-04	7	1		05/03/16 21:43
23	1LM.LM34708	WG567321-03 CCV (1.0ug/L)	1	1	STD75510	05/03/16 22:02
24	1LM.LM34709	WG567013-08 MRL (2.0ug/kg)	7	1	STD75510	05/03/16 22:21
25	1LM.LM34710	WG567321-04 CCB	1	1		05/03/16 22:40
26	1LM.LM34711	L16041363-05 (5x)	7	5		05/03/16 22:59
27	1LM.LM34712	L16041363-06	7	1		05/03/16 23:18
28	1LM.LM34713	L16041363-10 (5x)	7	5		05/03/16 23:37
29	1LM.LM34714	WG567321-05 CCV (1.0ug/L)	1	1	STD75510	05/03/16 23:56
30	1LM.LM34715	WG567013-09 MRL (2.0ug/kg)	7	1	STD75510	05/04/16 00:15
31	1LM.LM34716	WG567321-06 CCB	1	1		05/04/16 00:34

Comments

Page: 1 Approved: 05-MAY-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 050316_JWR.TXT

Analyst1: JWR Analyst2: NA

Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254

Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG567013 (soils)

Internal STD: COA18071 Surrogate STD: NA STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 STD75512

Comments

Seq.	Rerun	Dil.	Reason	Analytes
17				
	L16041363-08 MS : The MS %Rec is 129%. This is above the advisory limit of 120%. The parent sample to this MS had responses that passed the ion-ratio criteria, but had a quantified value below the method's detection limit, resulting in an assigned value of zero. If the quantified value for the parent sample were used in the %Rec calculation, the MS %Rec would be 93.9%.			
18				
	L16041363-09 MSD : The MSD %Rec is 131%. This is above the advisory limit of 120%. The parent sample to this MSD had responses that passed the ion-ratio criteria, but had a quantified value below the method's detection limit, resulting in an assigned value of zero. If the quantified value for the parent sample were used in the %Rec calculation, the MSD %Rec would be 95.4%.			

Page: 2 Approved: 05-MAY-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 062416_JWR.TXT
Analyst1: JWR Analyst2: NA
Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254
Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG574101 (waters)

Internal STD: STD18071 Surrogate STD: NA Calibration STD STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 MS/MSD STD: NA

Comments: Samples initially analyzed at dilutions were done so based on their historical results. Samples tagged with (NR) did not conform to their historical results. These samples were reanalyzed on the end of the sequence as neat or at dilutions based on the results of their initial analysis.

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	1LM.LM35462	WG574102-01 CCB	1	1		06/24/16 19:50
2	1LM.LM35463	WG574102-02 CCV (1.0ug/L)	1	1	STD75510	06/24/16 20:09
3	1LM.LM35464	WG574101-05 MRL (0.2ug/L)	1	1	STD75510	06/24/16 20:28
4	1LM.LM35465	WG574101-01 MCT (0.2ug/L)	1	1	STD75512	06/24/16 20:47
5	1LM.LM35466	WG574101-02 BLANK	1	1		06/24/16 21:06
6	1LM.LM35467	WG574101-03 LCS (0.2ug/L)	1	1	STD75512	06/24/16 21:25
7	1LM.LM35468	L16060823-01	1	1		06/24/16 21:44
8	1LM.LM35469	L16060823-03	1	1		06/24/16 22:03
9	1LM.LM35470	L16060823-04 (10x)	1	10		06/24/16 22:22
10	1LM.LM35471	L16060927-01 (100x) (NR)	1	100		06/24/16 22:41
11	1LM.LM35472	L16060927-02 (100x) (NR)	1	100		06/24/16 23:00
12	1LM.LM35473	L16060927-03	1	1		06/24/16 23:18
13	1LM.LM35474	L16060927-04	1	1		06/24/16 23:37
14	1LM.LM35475	WG574102-03 CCV (1.0ug/L)	1	1	STD75510	06/24/16 23:56
15	1LM.LM35476	WG574101-06 MRL (0.2ug/L)	1	1	STD75510	06/25/16 00:15
16	1LM.LM35477	WG574102-04 CCB	1	1		06/25/16 00:34
17	1LM.LM35478	L16061026-01 (100x) (NR)	1	100		06/25/16 00:53
18	1LM.LM35479	L16061026-03 (100x)	1	100		06/25/16 01:12
19	1LM.LM35480	L16061026-04 (10,000x)	1	10000		06/25/16 01:31
20	1LM.LM35481	L16061026-05 (100,000x) (NR)	1	100000		06/25/16 01:50
21	1LM.LM35482	L16061026-07 (10x) (NR)	1	10		06/25/16 02:09
22	1LM.LM35483	L16061142-01	1	1		06/25/16 02:28
23	1LM.LM35484	L16061142-02	1	1		06/25/16 02:47
24	1LM.LM35485	L16061142-04	1	1		06/25/16 03:06
25	1LM.LM35486	L16061142-06 (10x) (NR)	1	10		06/25/16 03:25
26	1LM.LM35487	L16061142-07	1	1		06/25/16 03:44
27	1LM.LM35488	WG574102-05 CCV (1.0ug/L)	1	1	STD75510	06/25/16 04:03
28	1LM.LM35489	WG574101-07 MRL (0.2ug/L)	1	1	STD75510	06/25/16 04:22
29	1LM.LM35490	WG574102-06 CCB	1	1		06/25/16 04:40
30	1LM.LM35491	WG574101-04 LCS2 (0.2ug/L)	1	1	STD75512	06/25/16 04:59
31	1LM.LM35492	L16061144-01	1	1		06/25/16 05:18
32	1LM.LM35493	L16061144-02	1	1		06/25/16 05:37
33	1LM.LM35494	L16061144-03 (10x) (NR)	1	10		06/25/16 05:56

Page: 1 Approved: 28-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 062416_JWR.TXT
Analyst1: JWR Analyst2: NA
Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254
Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG574101 (waters)

Internal STD: STD18071 Surrogate STD: NA STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 NA

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
34	1LM.LM35495	WG574102-07 CCV (1.0ug/L)	1	1	STD75510	06/25/16 06:15
35	1LM.LM35496	WG574101-08 MRL (0.2ug/L)	1	1	STD75510	06/25/16 06:34
36	1LM.LM35497	WG574102-08 CCB	1	1		06/25/16 06:53
37	1LM.LM35498	WG574102-09 CCV (1.0ug/L)	1	1	STD75510	06/25/16 18:06
38	1LM.LM35499	WG574101-09 MRL (0.2ug/L)	1	1	STD75510	06/25/16 18:25
39	1LM.LM35500	WG574102-10 CCB	1	1		06/25/16 18:44
40	1LM.LM35501	L16060927-01 (RR Neat)	1	1		06/25/16 19:02
41	1LM.LM35502	L16060927-02 (RR 1000x)	1	1000		06/25/16 19:21
42	1LM.LM35503	L16061026-01 (RR Neat)	1	1		06/25/16 19:40
43	1LM.LM35504	L16061026-05 (RR 10,000x)	1	10000		06/25/16 19:59
44	1LM.LM35505	L16061026-07 (RR 50x)	1	50		06/25/16 20:18
45	1LM.LM35506	L16061142-06 (RR Neat)	1	1		06/25/16 20:37
46	1LM.LM35507	L16061144-03 (RR Neat)	1	1		06/25/16 20:56
47	1LM.LM35508	WG574102-11 CCV (1.0ug/L)	1	1	STD75510	06/25/16 21:15
48	1LM.LM35509	WG574101-10 MRL (0.2ug/L)	1	1	STD75510	06/25/16 21:34
49	1LM.LM35510	WG574102-12 CCB	1	1		06/25/16 21:53

Comments

Seq.	Rerun	Dil.	Reason	Analytes

Page: 2 Approved: 28-JUN-16

Microbac Laboratories Inc.

Data Checklist

Date: 03-MAY-2016
Analyst: JWR
Analyst: NA
Method: 6850
Instrument: LCMS1
Curve Workgroup: WG567320
Runlog ID: 74891
Analytical Workgroups: L16041363 (SOILS)

ANALYTICAL	
System Performance Check	NA
DFTPP (GCMS)	NA
Endrin/DDT breakdown (8081/GCMS)	NA
Pentachlorophenol/benzidine tailing (GCMS)	NA
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	X
Average RF	NA
Linear regression or higher order curve	X
Alternate source standard (ICV) % Difference	X
Continuing Calibration (CCV)	X
% D/% Drift	X
Minimum response factors (GCMS)	X
Continuing calibration blank (CCB) (IC/LCMS)	X
Limit of quantitation verification (LOQV) (LCMS)	X
Special standards	NA
Blanks	X
TCL hits	ND
Surrogate recoveries	NA
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	NA
MS/MSD/Sample duplicates	X
Recoveries	X
%RPD	X
Interference check sample (ICS) (LCMS)	MCT
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	NA
Surrogate recoveries	NA
Internal standard areas (MS)	X
Library searches (GCMS)	NA
Calculations & correct factors	X
Compounds above calibration range	NA
Reruns	NA
Manual integrations	NA
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	
Check for completeness	X
Primary Reviewer	JWR
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	WTD

Primary Reviewer: 04-MAY-2016 John Richards

Secondary Reviewer: 05-MAY-2016 Wadei D

CHECKLIST1 - Modified 03/05/2008
Generated: MAY-05-2016 16:23:46

Microbac Laboratories Inc.

Data Checklist

Date: 24-JUN-2016
Analyst: JWR
Analyst: NA
Method: 6850
Instrument: LCMS1
Curve Workgroup: NA
Runlog ID: 75960
Analytical Workgroups: L16060823, 0927, 1026, 1142, 1144

ANALYTICAL	
System Performance Check	NA
DFTPP (GCMS)	NA
Endrin/DDT breakdown (8081/GCMS)	NA
Pentachlorophenol/benzidine tailing (GCMS)	NA
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	NA
Average RF	NA
Linear regression or higher order curve	NA
Alternate source standard (ICV) % Difference	NA
Continuing Calibration (CCV)	X
% D/% Drift	X
Minimum response factors (GCMS)	X
Continuing calibration blank (CCB) (IC/LCMS)	X
Limit of quantitation verification (LOQV) (LCMS)	X
Special standards	NA
Blanks	X
TCL hits	ND
Surrogate recoveries	NA
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	NA
MS/MSD/Sample duplicates	NA
Recoveries	NA
%RPD	NA
Interference check sample (ICS) (LCMS)	MCT
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	NA
Surrogate recoveries	NA
Internal standard areas (MS)	X
Library searches (GCMS)	NA
Calculations & correct factors	X
Compounds above calibration range	X
Reruns	X
Manual integrations	MCT-ONLY
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	
Check for completeness	X
Primary Reviewer	JWR
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	WTD

Primary Reviewer:
27-JUN-2016
John Richards

Secondary Reviewer:
28-JUN-2016
Wadei D

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-28-2016 14:00:55

Microbac Laboratories Inc.

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 6850 AAB#: WG574101

Login Number: L16060927

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18CPTMW23SW-061516	01	06/15/16					06/24/2016	9.5	28		06/25/16	1	28	
18CPTMW23-061516	02	06/15/16					06/24/2016	9.4	28		06/25/16	1	28	
18CPTMW19SW-061516	03	06/15/16					06/24/2016	9.4	28		06/24/16	.2	28	
18CPTMW19-061516	04	06/15/16					06/24/2016	9.3	28		06/24/16	.2	28	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4827770
Report generated 06/28/2016 14:16

METHOD BLANK SUMMARY

Login Number: L16060927
Blank File ID: 1LM.LM35466
Prep Date: 06/24/16 19:00
Analyzed Date: 06/24/16 21:06
Analyst: JWR

Work Group: WG574101
Blank Sample ID: WG574101-02
Instrument ID: LCMS1
Method: 6850

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
QCMRL	WG574101-05	1LM.LM35464	06/24/16 20:28	01
MCT	WG574101-01	1LM.LM35465	06/24/16 20:47	01
LCS	WG574101-03	1LM.LM35467	06/24/16 21:25	01
18CPTMW19SW-061516	L16060927-03	1LM.LM35473	06/24/16 23:18	01
18CPTMW19-061516	L16060927-04	1LM.LM35474	06/24/16 23:37	01
QCMRL	WG574101-06	1LM.LM35476	06/25/16 00:15	01
QCMRL	WG574101-07	1LM.LM35489	06/25/16 04:22	01
LCS2	WG574101-04	1LM.LM35491	06/25/16 04:59	01
QCMRL	WG574101-08	1LM.LM35496	06/25/16 06:34	01
QCMRL	WG574101-09	1LM.LM35499	06/25/16 18:25	01
18CPTMW23SW-061516	L16060927-01	1LM.LM35501	06/25/16 19:02	01
18CPTMW23-061516	L16060927-02	1LM.LM35502	06/25/16 19:21	DL01
QCMRL	WG574101-10	1LM.LM35509	06/25/16 21:34	01

Report Name: BLANK_SUMMARY
PDF File ID: 4827771
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16060927 Prep Date: 06/24/16 19:00 Sample ID: WG574101-02
Instrument ID: LCMS1 Run Date: 06/24/16 21:06 Prep Method: 6850
File ID: 1LM.LM35466 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Perchlorate	0.100	0.400	0.100	1	U

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name:BLANK
PDF ID: 4827772
28-JUN-2016 14:16

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16060927 Analyst: JWR Prep Method: 6850
Instrument ID: LCMS1 Matrix: Water Method: 6850
Workgroup (AAB#): WG574101 Units: ug/L
QC Key: DOD4 Lot #: STD75512

Sample ID: WG574101-03 LCS File ID: 1LM.LM35467 Run Date: 06/24/2016 21:25
Sample ID: WG574101-04 LCS2 File ID: 1LM.LM35491 Run Date: 06/25/2016 04:59

Analytes	LCS			LCS2			%RPD	%Rec Limits	RPD Lmt	Q
	Known	Found	% REC	Known	Found	% REC				
Perchlorate	0.200	0.208	104	0.200	0.213	107	2.38	80 - 120	15	

LCS_LCS2 - Modified 03/06/2008
PDF File ID: 4827773
Report generated: 06/28/2016 14:16

Microbac Laboratories Inc.

INITIAL CALIBRATION SUMMARY

Login Number: L16060927
Analytical Method: 6850
ICAL Workgroup: WG567320

Instrument ID: LCMS1
Initial Calibration Date: 03-MAY-16 17:18
Column ID: F

Analyte		AVG RF	% RSD	LINEAR (R)	QUAD (R^2)
Perchlorate		1.699	4.81	1.00000	

R = Correlation coefficient; 0.995 minimum
R^2 = Coefficient of determination; 0.99 minimum

INT_CAL - Modified 03/06/2008
PDF File ID: 4828715
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16060927

Analytical Method: 6850

Instrument ID: LCMS1

Initial Calibration Date: 03-MAY-16 17:18

Column ID: F

Analyte	WG567320-02			WG567320-03			WG567320-04		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
Perchlorate	0.100	17900.0000	1.792	0.200	34100.0000	1.718	0.500	82200.0000	1.637

INT_CAL - Modified 03/06/2008
PDF File ID: 4828715
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16060927

Analytical Method: 6850

Instrument ID: LCMS1

Initial Calibration Date: 03-MAY-16 17:18

Column ID: F

Analyte	WG567320-05			WG567320-06			WG567320-07		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
Perchlorate	1.00	168000.000	1.697	2.00	330000.000	1.672	5.00	810000.000	1.695

INT_CAL - Modified 03/06/2008
PDF File ID: 4828715
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16060927

Analytical Method: 6850

Instrument ID: LCMS1

Initial Calibration Date: 03-MAY-16 17:18

Column ID: F

Analyte	WG567320-08		
	CONC	RESP	RF
Perchlorate	10.0	1530000.00	1.680

INT_CAL - Modified 03/06/2008

PDF File ID: 4828715

Report generated 06/28/2016 14:16

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16060927 Run Date: 05/03/2016 Sample ID: WG567320-09
Instrument ID: LCMS1 Run Time: 17:37 Method: 6850
File ID: 1LM.LM34694 Analyst: JWR QC Key: DOD4
ICal Workgroup: WG567320 Cal ID: LCMS1 - 03-MAY-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
Perchlorate		1.00	0.985	ug/L	1.66	1.50	15	

* Exceeds %D Limit

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4828716
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060927
Run Date: 06/24/2016
Sample ID: WG574102-01
Instrument ID: LCMS1
Run Time: 19:50
Method: 6850
File ID: 1LM.LM35462
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827776
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060927
Run Date: 06/25/2016
Sample ID: WG574102-04
Instrument ID: LCMS1
Run Time: 00:34
Method: 6850
File ID: 1LM.LM35477
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827776
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060927
Run Date: 06/25/2016
Sample ID: WG574102-06
Instrument ID: LCMS1
Run Time: 04:40
Method: 6850
File ID: 1LM.LM35490
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827776
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060927
Run Date: 06/25/2016
Sample ID: WG574102-08
Instrument ID: LCMS1
Run Time: 06:53
Method: 6850
File ID: 1LM.LM35497
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827776
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060927
Run Date: 06/25/2016
Sample ID: WG574102-10
Instrument ID: LCMS1
Run Time: 18:44
Method: 6850
File ID: 1LM.LM35500
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827776
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16060927
Run Date: 06/25/2016
Sample ID: WG574102-12
Instrument ID: LCMS1
Run Time: 21:53
Method: 6850
File ID: 1LM.LM35510
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827776
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927
Run Date: 06/24/2016
Sample ID: WG574102-02
Instrument ID: LCMS1
Run Time: 20:09
Method: 6850
File ID: 1LM.LM35463
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.01	ug/L	1.70	1.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827775
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927
Run Date: 06/24/2016
Sample ID: WG574102-03
Instrument ID: LCMS1
Run Time: 23:56
Method: 6850
File ID: 1LM.LM35475
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.01	ug/L	1.70	1.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827775
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927
Run Date: 06/25/2016
Sample ID: WG574102-05
Instrument ID: LCMS1
Run Time: 04:03
Method: 6850
File ID: 1LM.LM35488
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.04	ug/L	1.75	4.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827775
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927
Run Date: 06/25/2016
Sample ID: WG574102-07
Instrument ID: LCMS1
Run Time: 06:15
Method: 6850
File ID: 1LM.LM35495
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.02	ug/L	1.72	2.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827775
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927
Run Date: 06/25/2016
Sample ID: WG574102-09
Instrument ID: LCMS1
Run Time: 18:06
Method: 6850
File ID: 1LM.LM35498
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.04	ug/L	1.76	4.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827775
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16060927 Run Date: 06/25/2016 Sample ID: WG574102-11
Instrument ID: LCMS1 Run Time: 21:15 Method: 6850
File ID: 1LM.LM35508 Analyst: JWR QC Key: DOD4
Workgroup (AAB#): WG574101 Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.02	ug/L	1.73	2.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827775
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060927 Run Date: 06/24/2016 Sample ID: WG574101-05
Instrument ID: LCMS1 Run Time: 20:28 Prep Method: 6850
File ID: 1LM.LM35464 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.206	103	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827774
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060927 Run Date: 06/25/2016 Sample ID: WG574101-06
Instrument ID: LCMS1 Run Time: 00:15 Prep Method: 6850
File ID: 1LM.LM35476 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.198	99.0	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827774
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060927 Run Date: 06/25/2016 Sample ID: WG574101-07
Instrument ID: LCMS1 Run Time: 04:22 Prep Method: 6850
File ID: 1LM.LM35489 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.211	106	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827774
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060927 Run Date: 06/25/2016 Sample ID: WG574101-08
Instrument ID: LCMS1 Run Time: 06:34 Prep Method: 6850
File ID: 1LM.LM35496 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.201	101	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827774
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060927 Run Date: 06/25/2016 Sample ID: WG574101-09
Instrument ID: LCMS1 Run Time: 18:25 Prep Method: 6850
File ID: 1LM.LM35499 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.207	104	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827774
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16060927 Run Date: 06/25/2016 Sample ID: WG574101-10
Instrument ID: LCMS1 Run Time: 21:34 Prep Method: 6850
File ID: 1LM.LM35509 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.206	103	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827774
Report generated 06/28/2016 14:16

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO AVERAGE OF ICAL)

Login Number: L16060927
Instrument ID: LCMS1
Workgroup (AAB#): WG574101

ICAL CCV Number: WG567320-05
CAL ID: LCMS1-03-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1
WG567320	NA	NA	489000
Upper Limit	NA	NA	733500
Lower Limit	NA	NA	244500
L16060927-01	1.00	01	390000
L16060927-02	1000	DL01	471000
L16060927-03	1.00	01	418000
L16060927-04	1.00	01	463000
WG574101-02	1.00	01	443000
WG574101-03	1.00	01	444000
WG574101-04	1.00	01	526000

IS-1 - O18LP

Underline = Response outside limits

INTERNAL_STD_AVG_ICAL - Modified 03/10/2010
PDF File ID: 4827777
Report generated 06/28/2016 14:16

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	L16060927-01
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35501
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 19:02	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	53800	19600	2.74	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	L16060927-02
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35502
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 19:21	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	367000	129000	2.84	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	L16060927-03
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35473
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 23:18	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	176000	61700	2.85	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	L16060927-04
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35474
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 23:37	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	828000	283000	2.93	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG567320-02
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34687
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 15:25	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	17900	6950	2.58	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG567320-03
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34688
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 15:43	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	34100	11900	2.87	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG567320-04
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34689
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:02	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	82200	29400	2.80	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG567320-05
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34690
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:21	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	168000	56600	2.97	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG567320-06
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34691
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:40	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	330000	108000	3.06	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG567320-07
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34692
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:59	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	810000	269000	3.01	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG567320-08
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34693
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 17:18	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	1530000	512000	2.99	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG567320-09
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34694
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 17:37	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	169000	56300	3.00	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	WG574101-01
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35465
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 20:47	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	28100	10000	2.81	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	WG574101-02
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35466
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 21:06	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	WG574101-03
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35467
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 21:25	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	31600	11800	2.68	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	WG574101-04
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35491
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:59	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	38300	14000	2.74	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	WG574101-05
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35464
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 20:28	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	31600	12100	2.61	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	WG574101-06
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35476
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 00:15	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	35000	13100	2.67	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	WG574101-07
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35489
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:22	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	38900	13400	2.90	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	WG574101-08
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35496
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 06:34	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	38000	14100	2.70	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	WG574101-09
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35499
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 18:25	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	32900	11500	2.86	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**	6850	**Samplenum:**	WG574101-10
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35509
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 21:34	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	37900	14000	2.71	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-01
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:50	**File ID:**	1LM.LM35462
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 19:50	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-02
Instrument:	LCMS1	**Prep Date:**	06/24/2016 20:09	**File ID:**	1LM.LM35463
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 20:09	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	152000	55500	2.74	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-03
Instrument:	LCMS1	**Prep Date:**	06/24/2016 23:56	**File ID:**	1LM.LM35475
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 23:56	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	170000	57500	2.96	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-04
Instrument:	LCMS1	**Prep Date:**	06/25/2016 00:34	**File ID:**	1LM.LM35477
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 00:34	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	824	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-05
Instrument:	LCMS1	**Prep Date:**	06/25/2016 04:03	**File ID:**	1LM.LM35488
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:03	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	179000	60200	2.97	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-06
Instrument:	LCMS1	**Prep Date:**	06/25/2016 04:40	**File ID:**	1LM.LM35490
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:40	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-07
Instrument:	LCMS1	**Prep Date:**	06/25/2016 06:15	**File ID:**	1LM.LM35495
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 06:15	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	189000	63400	2.98	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-08
Instrument:	LCMS1	**Prep Date:**	06/25/2016 06:53	**File ID:**	1LM.LM35497
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 06:53	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	518	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-09
Instrument:	LCMS1	**Prep Date:**	06/25/2016 18:06	**File ID:**	1LM.LM35498
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 18:06	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	161000	56700	2.84	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-10
Instrument:	LCMS1	**Prep Date:**	06/25/2016 18:44	**File ID:**	1LM.LM35500
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 18:44	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-11
Instrument:	LCMS1	**Prep Date:**	06/25/2016 21:15	**File ID:**	1LM.LM35508
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 21:15	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	176000	59300	2.97	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16060927	**Prep Method:**		**Samplenum:**	WG574102-12
Instrument:	LCMS1	**Prep Date:**	06/25/2016 21:53	**File ID:**	1LM.LM35510
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 21:53	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

2.2.2.3 Sample Data

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Data File	LM35501.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 7:02:57 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16060927-01 (RR Neat)	**Injection Vial**	28.00
Data File	LM35501.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 7:02:57 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16060927-01	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	3.900e+05	9.36	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	5.380e+04	9.37	N/A	0.407
Perchlorate conf	1.960e+04	9.36	N/A	0.431

O18LP (Internal Standard)

RT (Exp. RT): 9.36(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.37 (9.84) min
Calculated conc:	0.407 ng/ml
Area Ratio:	0.138
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.36 (9.84) min
Calculated conc:	0.431 ng/ml
Area Ratio:	0.05
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:53 PM

Data File	LM35502.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 7:21:52 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16060927-02 (RR 1000x)	**Injection Vial**	29.00
Data File	LM35502.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 7:21:52 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16060927-02	**Dilution Factor**	1.00
Sample Comment	11,000.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.710e+05	9.62	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.670e+05	9.63	N/A	2.31
Perchlorate conf	1.290e+05	9.62	N/A	2.44

O18LP (Internal Standard)

RT (Exp. RT): 9.62(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.63 (9.84) min
Calculated conc:	2.31 ng/ml
Area Ratio:	0.779
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.62 (9.84) min
Calculated conc:	2.44 ng/ml
Area Ratio:	0.274
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:42 PM

Data File	LM35473.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 11:18:58 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16060927-03	**Injection Vial**	13.00
Data File	LM35473.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 11:18:58 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16060927-03	**Dilution Factor**	1.00
Sample Comment	1,1 (Hist)	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.180e+05	9.37	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.760e+05	9.38	N/A	1.25
Perchlorate conf	6.170e+04	9.37	N/A	1.30

O18LP (Internal Standard)

RT (Exp. RT): 9.37(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:42 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.38 (9.84) min
Calculated conc:	1.25 ng/ml
Area Ratio:	0.421
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.37 (9.84) min
Calculated conc:	1.30 ng/ml
Area Ratio:	0.147
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:42 PM

Data File	LM35474.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 11:37:53 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16060927-04	**Injection Vial**	14.00
Data File	LM35474.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 11:37:53 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16060927-04	**Dilution Factor**	1.00
Sample Comment	1,1 (Hist)	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.630e+05	9.50	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	8.280e+05	9.51	N/A	5.31
Perchlorate conf	2.830e+05	9.50	N/A	5.46

O18LP (Internal Standard)

RT (Exp. RT): 9.50(9.84) min
Concentration: 5.00 ug/L
Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:42 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.51 (9.84) min
Calculated conc:	5.31 ng/ml
Area Ratio:	1.787
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.50 (9.84) min
Calculated conc:	5.46 ng/ml
Area Ratio:	0.612
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

2.2.2.4 Standards Data

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Data File	LM34686.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-01 CCB	**Injection Vial**	1.00
Data File	LM34686.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-01	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.020e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min
Concentration: 5.00 ug/L
Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:20 PM

Analyte Name: **Perchlorate**
Internal Standard: O18LP

Data File	LM34686.wiff	**Result Table**	050316_JWR.rdb
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Regression Equation: y = 1.68 x + 0.00128 (r = 1.0000)

Expected Concentration	Number of Values	Mean Calculated Concentration	% Accuracy	Std. Deviation	%CV
0.10	1	0.10	102.8	N/A	N/A
0.20	1	0.20	100.3	N/A	N/A
0.50	1	0.48	96.6	N/A	N/A
1.00	1	1.01	100.5	N/A	N/A
2.00	1	1.99	99.3	N/A	N/A
5.00	1	5.04	100.7	N/A	N/A
10.00	1	9.99	99.9	N/A	N/A



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:20 PM

Analyte Name: **Perchlorate conf**
Internal Standard: O18LP

Data File	LM34686.wiff	**Result Table**	050316_JWR.rdb
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Regression Equation: y = 0.558 x + 0.00228 (r = 0.9999)

Expected Concentration	Number of Values	Mean Calculated Concentration	% Accuracy	Std. Deviation	%CV
0.10	1	0.10	104.3	N/A	N/A
0.20	1	0.19	96.8	N/A	N/A
0.50	1	0.50	100.6	N/A	N/A
1.00	1	1.00	100.5	N/A	N/A
2.00	1	1.94	97.2	N/A	N/A
5.00	1	5.02	100.4	N/A	N/A
10.00	1	10.03	100.3	N/A	N/A



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Data File	LM34687.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 3:25:04 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-02 STD (0.1 ug/L)	**Injection Vial**	2.00
Data File	LM34687.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 3:25:04 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-02	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.000e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.790e+04	9.85	0.10	0.103
Perchlorate conf	6.950e+03	9.83	0.10	0.104

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.103 ng/ml
Area Ratio:	0.036
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	0.104 ng/ml
Area Ratio:	0.014
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34688.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 3:43:59 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-03 STD (0.2 ug/L)	**Injection Vial**	3.00
Data File	LM34688.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 3:43:59 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-03	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.970e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.410e+04	9.85	0.20	0.201
Perchlorate conf	1.190e+04	9.84	0.20	0.194

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.201 ng/ml
Area Ratio:	0.069
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	0.194 ng/ml
Area Ratio:	0.024
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34689.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:02:52 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-04 STD (0.5 ug/L)	**Injection Vial**	4.00
Data File	LM34689.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:02:52 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-04	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.020e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	8.220e+04	9.85	0.50	0.483
Perchlorate conf	2.940e+04	9.84	0.50	0.503

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.483 ng/ml
Area Ratio:	0.164
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	0.503 ng/ml
Area Ratio:	0.058
Sample Type:	(Standard)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34690.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:21:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-05 STD (1.0 ug/L)	**Injection Vial**	5.00
Data File	LM34690.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:21:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-05	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.950e+05	9.83	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.680e+05	9.84	1.00	1.01
Perchlorate conf	5.660e+04	9.83	1.00	1.00

O18LP (Internal Standard)

RT (Exp. RT): 9.83(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.339
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	1.00 ng/ml
Area Ratio:	0.114
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34691.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:40:45 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-06 STD (2.0 ug/L)	**Injection Vial**	6.00
Data File	LM34691.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:40:45 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-06	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.930e+05	9.83	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.300e+05	9.84	2.00	1.99
Perchlorate conf	1.080e+05	9.83	2.00	1.94

O18LP (Internal Standard)

RT (Exp. RT): 9.83(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	1.99 ng/ml
Area Ratio:	0.669
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	1.94 ng/ml
Area Ratio:	0.219
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Data File	LM34692.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:59:42 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-07 STD (5.0 ug/L)	**Injection Vial**	7.00
Data File	LM34692.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:59:42 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-07	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.780e+05	9.83	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	8.100e+05	9.84	5.00	5.04
Perchlorate conf	2.690e+05	9.83	5.00	5.02

O18LP (Internal Standard)

RT (Exp. RT): 9.83(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	5.04 ng/ml
Area Ratio:	1.695
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	5.02 ng/ml
Area Ratio:	0.562
Sample Type:	(Standard)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Data File	LM34693.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 5:18:37 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-08 STD (10 ug/L)	**Injection Vial**	8.00
Data File	LM34693.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 5:18:37 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-08	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.560e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.530e+06	9.85	10.00	9.99
Perchlorate conf	5.120e+05	9.83	10.00	10.00

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	9.99 ng/ml
Area Ratio:	3.36
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	10.00 ng/ml
Area Ratio:	1.122
Sample Type:	(Standard)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:23 PM

Data File	LM34694.wiff	**Result Table**	050316_JWR.rdb
Acquisition Date	5/3/2016 5:37:34 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-09 SSCV (1.0 ug/L)	**Injection Vial**	9.00
Data File	LM34694.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 5:37:34 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	050316_JWR.rdb
Sample ID	WG567320-09	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.080e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.690e+05	9.85	1.00	0.985
Perchlorate conf	5.630e+04	9.84	1.00	0.973

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:23 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.985 ng/ml
Area Ratio:	0.332
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	0.973 ng/ml
Area Ratio:	0.111
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Data File	LM35463.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 8:09:38 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-02 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35463.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 8:09:38 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-02	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.450e+05	9.34	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.520e+05	9.35	1.00	1.01
Perchlorate conf	5.550e+04	9.33	1.00	1.10

O18LP (Internal Standard)

RT (Exp. RT): 9.34(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.341
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.33 (9.84) min
Calculated conc:	1.10 ng/ml
Area Ratio:	0.125
Sample Type:	(Quality Control)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Data File	LM35475.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 11:56:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-03 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35475.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 11:56:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-03	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.980e+05	9.32	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.700e+05	9.33	1.00	1.01
Perchlorate conf	5.750e+04	9.32	1.00	1.01

O18LP (Internal Standard)

RT (Exp. RT): 9.32(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.33 (9.84) min Calculated conc: 1.01 ng/ml Area Ratio: 0.34 Sample Type: (Quality Control)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.32 (9.84) min Calculated conc: 1.01 ng/ml Area Ratio: 0.115 Sample Type: (Quality Control)	

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:47 PM

Data File	LM35488.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:03:05 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-05 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35488.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:03:05 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-05	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.110e+05	9.33	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.790e+05	9.34	1.00	1.04
Perchlorate conf	6.020e+04	9.32	1.00	1.03

O18LP (Internal Standard)

RT (Exp. RT): 9.33(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.34 (9.84) min
Calculated conc:	1.04 ng/ml
Area Ratio:	0.35
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.32 (9.84) min
Calculated conc:	1.03 ng/ml
Area Ratio:	0.118
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:50 PM

Data File	LM35495.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:15:37 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-07 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35495.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:15:37 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-07	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.500e+05	9.33	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.890e+05	9.35	1.00	1.02
Perchlorate conf	6.340e+04	9.33	1.00	1.01

O18LP (Internal Standard)

RT (Exp. RT): 9.33(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:50 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.02 ng/ml
Area Ratio:	0.345
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.33 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.115
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Data File	LM35498.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:06:10 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-09 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35498.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:06:10 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-09	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.580e+05	9.34	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.610e+05	9.35	1.00	1.04
Perchlorate conf	5.670e+04	9.34	1.00	1.09

O18LP (Internal Standard)

RT (Exp. RT): 9.34(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.04 ng/ml
Area Ratio:	0.353
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.34 (9.84) min
Calculated conc:	1.09 ng/ml
Area Ratio:	0.124
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:55 PM

Data File	LM35508.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 9:15:29 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-11 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35508.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 9:15:29 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-11	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.090e+05	9.35	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.760e+05	9.36	1.00	1.02
Perchlorate conf	5.930e+04	9.35	1.00	1.02

O18LP (Internal Standard)

RT (Exp. RT): 9.35(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.36 (9.84) min
Calculated conc:	1.02 ng/ml
Area Ratio:	0.345
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.02 ng/ml
Area Ratio:	0.117
Sample Type:	(Quality Control)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Data File	LM35464.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 8:28:33 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-05 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35464.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 8:28:33 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-05	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.490e+05	9.39	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.160e+04	9.40	N/A	0.206
Perchlorate conf	1.210e+04	9.37	N/A	0.222

O18LP (Internal Standard)

RT (Exp. RT): 9.39(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.40 (9.84) min
Calculated conc:	0.206 ng/ml
Area Ratio:	0.07
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.37 (9.84) min
Calculated conc:	0.222 ng/ml
Area Ratio:	0.027
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Data File	LM35476.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 12:15:45 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-06 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35476.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 12:15:45 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-06	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.160e+05	9.36	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.500e+04	9.37	N/A	0.198
Perchlorate conf	1.310e+04	9.35	N/A	0.207

O18LP (Internal Standard)

RT (Exp. RT): 9.36(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.37 (9.84) min
Calculated conc:	0.198 ng/ml
Area Ratio:	0.068
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	0.207 ng/ml
Area Ratio:	0.025
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Data File	LM35489.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:22:00 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-07 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35489.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:22:00 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-07	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.380e+05	9.37	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.890e+04	9.39	N/A	0.211
Perchlorate conf	1.340e+04	9.38	N/A	0.202

O18LP (Internal Standard)

RT (Exp. RT): 9.37(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.39 (9.84) min
Calculated conc:	0.211 ng/ml
Area Ratio:	0.072
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.38 (9.84) min
Calculated conc:	0.202 ng/ml
Area Ratio:	0.025
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:50 PM

Data File	LM35496.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:34:33 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-08 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35496.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:34:33 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-08	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.520e+05	9.38	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.800e+04	9.39	N/A	0.201
Perchlorate conf	1.410e+04	9.38	N/A	0.208

O18LP (Internal Standard)

RT (Exp. RT): 9.38(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.39 (9.84) min
Calculated conc:	0.201 ng/ml
Area Ratio:	0.069
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.38 (9.84) min
Calculated conc:	0.208 ng/ml
Area Ratio:	0.026
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Data File	LM35499.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:25:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-09 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35499.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:25:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-09	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.640e+05	9.39	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.290e+04	9.40	N/A	0.207
Perchlorate conf	1.150e+04	9.38	N/A	0.201

O18LP (Internal Standard)

RT (Exp. RT): 9.39(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.40 (9.84) min
Calculated conc:	0.207 ng/ml
Area Ratio:	0.071
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.38 (9.84) min
Calculated conc:	0.201 ng/ml
Area Ratio:	0.025
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:55 PM

Data File	LM35509.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 9:34:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-10 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35509.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 9:34:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-10	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.370e+05	9.39	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.790e+04	9.40	N/A	0.206
Perchlorate conf	1.400e+04	9.39	N/A	0.214

O18LP (Internal Standard)

RT (Exp. RT): 9.39(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.40 (9.84) min
Calculated conc:	0.206 ng/ml
Area Ratio:	0.071
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.39 (9.84) min
Calculated conc:	0.214 ng/ml
Area Ratio:	0.026
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Data File	LM35462.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 7:50:39 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-01 CCB	**Injection Vial**	1.00
Data File	LM35462.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 7:50:39 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-01	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.420e+05	9.58	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.58(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Data File	LM35477.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 12:34:42 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-04 CCB	**Injection Vial**	1.00
Data File	LM35477.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 12:34:42 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-04	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.240e+05	9.57	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	8.240e+02	9.53	N/A	< 0

O18LP (Internal Standard)

RT (Exp. RT): 9.57(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.53 (9.84) min Calculated conc: < 0 ng/ml Area Ratio: 0.002 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Data File	LM35490.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:40:56 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-06 CCB	**Injection Vial**	1.00
Data File	LM35490.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:40:56 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-06	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.420e+05	9.59	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.59(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Data File	LM35497.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:53:28 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-08 CCB	**Injection Vial**	1.00
Data File	LM35497.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:53:28 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-08	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.550e+05	9.59	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	5.180e+02	9.63	N/A	< 0

O18LP (Internal Standard)

RT (Exp. RT): 9.59(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.63 (9.84) min Calculated conc: < 0 ng/ml Area Ratio: 0.001 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Data File	LM35500.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:44:00 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-10 CCB	**Injection Vial**	1.00
Data File	LM35500.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:44:00 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-10	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.840e+05	9.61	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.61(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:56 PM

Data File	LM35510.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 9:53:20 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-12 CCB	**Injection Vial**	1.00
Data File	LM35510.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 9:53:20 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-12	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.380e+05	9.61	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.61(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Data File	LM35465.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 8:47:28 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-01 MCT (0.2ug/L)	**Injection Vial**	4.00
Data File	LM35465.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 8:47:28 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-01	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.320e+05	8.95	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	2.810e+04	8.97	N/A	0.19
Perchlorate conf	1.000e+04	8.95	N/A	0.188

O18LP (Internal Standard)

RT (Exp. RT): 8.95(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	8.97 (9.84) min
Calculated conc:	0.19 ng/ml
Area Ratio:	0.065
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	8.95 (9.84) min
Calculated conc:	0.188 ng/ml
Area Ratio:	0.023
Sample Type:	(Unknown)



Printing Time: 4:34:28 PM
Printing Date: Saturday, June 25, 2016

Sample Name: "WG574101-01 MCT (0.2ug/L)" Sample ID: "WG574101-01" File: "LM35465.wiff"
Peak Name: "Perchlorate" Mass(es): "98.8/83.3 amu"
Comment: "1,1 STD75512" Annotation: ""

Sample Index:	1
Sample Type:	Unknown
Concentration:	N/A
Calculated Conc:	0.219 ug/L
Acq. Date:	6/24/2016
Acq. Time:	8:47:28 PM
Modified:	No
Proc. Algorithm:	Analyst Classic
Bunching Factor:	2
Noise Threshold:	18.48 cps
Area Threshold:	92.40 cps
Num. Smooths:	5
Exp. Width:	0.20
Sep. Height:	0.01
Exp. Peak Ratio:	5.00
Exp. Adj. Ratio:	4.00
Exp. Val. Ratio:	3.00
RT Window:	30.000 sec
Expected RT:	9.84 min
Use Relative RT:	Yes
Int. Type:	Valley
Retention Time:	8.97 min
Area:	3.23e+004 counts
Height:	1951.523 cps
Start Time:	8.526 min
End Time:	9.417 min



Collected by: N/A
Electronic Signature: no
Operator: lcmsl

Printing Time: 4:36:56 PM
Printing Date: Saturday, June 25, 2016

Sample Name: "WG574101-01 MCT (0.2ug/L)" Sample ID: "WG574101-01" File: "LM35465.wiff"
Peak Name: "Perchlorate" Mass(es): "98.8/83.3 amu"
Comment: "1,1 STD75512" Annotation: ""

Sample Index:	1
Sample Type:	Unknown
Concentration:	N/A
Calculated Conc:	0.190 ug/L
Acq. Date:	6/24/2016
Acq. Time:	8:47:28 PM
Modified:	Yes
Proc. Algorithm:	Analyst Classic
Smoothing Factor:	2
Noise Threshold:	18.48 cps
Area Threshold:	92.40 cps
Num. Smooths:	5
Sep. Width:	0.20
Sep. Height:	0.01
Exp. Peak Ratio:	5.00
Exp. Adj. Ratio:	4.00
Exp. Val. Ratio:	3.00
RT Window:	30.000 sec
Expected RT:	9.84 min
Use Relative RT:	Yes
Int. Type:	Manual
Retention Time:	8.97 min
Area:	2.81e+004 counts
Height:	1872.249 cps
Start Time:	8.587 min
End Time:	9.417 min



#4
JWR 06/25/16

✓ WTD
06-28-16

Collected by: N/A
Electronic Signature: no
Operator: lcms1

2.2.2.5 Raw QC Data

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Data File	LM35466.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 9:06:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-02 BLANK	**Injection Vial**	5.00
Data File	LM35466.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 9:06:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-02	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.430e+05	9.50	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.50(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Data File	LM35467.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 9:25:22 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-03 LCS (0.2ug/L)	**Injection Vial**	6.00
Data File	LM35467.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 9:25:22 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-03	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.440e+05	9.53	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.160e+04	9.54	N/A	0.208
Perchlorate conf	1.180e+04	9.53	N/A	0.217

O18LP (Internal Standard)

RT (Exp. RT): 9.53(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.54 (9.84) min
Calculated conc:	0.208 ng/ml
Area Ratio:	0.071
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.53 (9.84) min
Calculated conc:	0.217 ng/ml
Area Ratio:	0.027
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Data File	LM35491.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:59:54 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-04 LCS2 (0.2ug/L)	**Injection Vial**	7.00
Data File	LM35491.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:59:54 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-04	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.260e+05	9.57	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.830e+04	9.59	N/A	0.213
Perchlorate conf	1.400e+04	9.57	N/A	0.218

O18LP (Internal Standard)

RT (Exp. RT): 9.57(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.59 (9.84) min
Calculated conc:	0.213 ng/ml
Area Ratio:	0.073
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.57 (9.84) min
Calculated conc:	0.218 ng/ml
Area Ratio:	0.027
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

3.0 Attachments

Microbac Laboratories Inc.
Ohio Valley Division Analyst List
July 6, 2016

001 - BIO-CHEM TESTING WVDEP 220
002 - REIC Consultants, Inc. WVDEP 060
003 - Sturm Environmental
004 - MICROBAC PITTSBURGH
005 - ES LABORATORIES
006 - ALCOSAN LABORATORIES
007 - ALS LABORATORIES
008 - BENCHMARK LABORATORIES
010 - MICROBAC CHICAGOLAND
AC - AMBER R. CARMICHAEL
ADC - ANTHONY D. CANTER
ADG - APRIL D. GREENE
AED - ALLEN E. DAVIS
ALS - ADRIANE L. STEED
AMA - ALEXANDRA M. ALFRED
AWE - ANDREW W. ESSIG
AZH - AFTER HOURS
BJO - BRIAN J. OGDEN
BKT - BRENDAN TORRENCE
BLG - BRENDA L. GREENWALT
BNB - Brandi N. Bentley
BRG - BRENDA R. GREGORY
CAA - CASSIE A. AUGENSTEIN
CAF - CHERYL A. FLOWERS
CEB - CHAD E. BARNES
CJQ - Cameron J. Quick
CJR - COURTNEY J. REXROAD
CLC - CHRYS L. CRAWFORD
CLS - CARA L. STRICKLER
CLW - CHARISSA L. WINTERS
CPD - CHAD P. DAVIS
CSH - CHRIS S. HILL
CV - Carl Volkman
DAK - DEAN A. KETELSEN
DCM - DAVID C. MERCKLE
DEV - DAVID E. VANDENBERG
DIH - DEANNA I. HESSON
DLB - DAVID L. BUMGARNER
DLP - DOROTHY L. PAYNE
DLW - DIANA L. WRIGHT
DSM - DAVID S. MOSSOR
ECL - ERIC C. LAWSON
EMW - ERIC M. WILKEN
ENY - EMILY N. YOAK
ERP - ERIN R. PORTER
FJB - FRANCES J. BOLDEN
JBK - JEREMY B. KINNEY
JDH - JUSTIN D. HESSON
JDS - JARED D. SMITH
JJS - JOHN J. STE MARIE
JKP - JACQUELINE K. PARSONS
JLD - JESSICA L. DELONG
JLL - JOHN L. LENT
JMW - JEANA M. WHITE
JTP - JOSHUA T. PEMBERTON
JWR - JOHN W. RICHARDS
JWS - JACK W. SHEAVES
JYH - JI Y. HU
KAJ - KELLIE A. JOHNSON
KAT - KATHY A. TUCKER
KDW - KATHRYN D. WELCH
KEB - KATIE E. BARNES
KHR - KIM H. RHODES
KKB - KERRI K. BUCK
KRA - KATHY R. ALBERTSON
KRB - KAELY R. BECKER
KRP - KATHY R. PARSONS
LEC - LAURA E. CARPENTER
LJH - Lacey J. Hendershot
LKN - LINDA K. NEDEFF
LLS - LARRY L. STEPHENS
LSB - LESLIE S. BUCINA
MAP - MARLA A. PORTER
MBK - MORGAN B. KNOWLTON
MDA - MIKE D. ALBERTSON
MDC - MIKE D. COCHRAN
MES - MARY E. SCHILLING
MLB - MEGAN L. BACHE
MMB - MAREN M. BEERY
MRT - MICHELLE R. TAYLOR
MSW - MATT S. WILSON
NPH - Natalie P. Hart
PDM - PIERCE D. MORRIS
PIT - MICROBAC WARRENDALE
PRL - PAIGE R. LAMB
PSW - PEGGY S. WEBB
QX - QIN XU
RAH - ROY A. HALSTEAD
REK - BOB E. KYER
RLB - BOB BUCHANAN
RM - RAYMOND MALEKE
RNP - RICK N. PETTY
RST - ROBIN S. TURNER
SAV - SARAH A. VANDENBERG
SCB - SARAH C. BOGOLIN
SDC - SHALYN D. CONLEY
SLM - STEPHANIE L. MOSSBURG
SLP - SHERI L. PFALZGRAF
TB - TODD BOYLE
TGF - TIM G. FELTON
TMB - TIFFANY M. BAILEY
TMM - TAMMY M. MORRIS
VC - VICKI COLLIER
WJB - WILL J. BEASLEY
WRR - WESLEY R. RICHARDS
WTD - WADE T. DELONG
XXX - UNAVAILABLE OR SUBCONTRACT

Microbac Laboratories Inc.
List of Valid Qualifiers
July 06, 2016

Qualkey: DOD

Qualifier	Description
*	Surrogate or spike compound out of range
+	Correlation coefficient for the MSA is less than 0.995
<	Result is less than the associated numerical value.
>	Greater than
A	See the report narrative
B	The reported result is associated with a contaminated method blank.
B,H1	Analyte present in method blank. Sample analysis performed past holding time.
B1	Target analyte detected in method blank at or above the method reporting limit
B3	Target analyte detected in calibration blank at or above the method reporting limit
B4	The BOD unseeded dilution water blank exceeded 0.2 mg/L
C	Confirmed by GC/MS
CG	Confluent growth
CT1	Cooler temperature at sample reciept exceeded regulatory limit.
DL	Surrogate or spike compound was diluted out
E	Estimated concentration due to sample matrix interference
E,CT1	Estimated results. The cooler temperature at receipt exceeded regulatory guidelines for requested testing.
EDL	Elevated sample reporting limits, presence of non-target analytes
EMPC	Estimated Maximum Possible Concentration
F, S	Estimated result below quantitation limit; method of standard additions(MSA)
F,CT1	Estimated value; the analyte concentration was less than the RL/LOQ. The cooler temperature at receipt exceeded regula
FL	Free Liquid
FP1	Did not ignite.
H1	Sample analysis performed past holding time.
H1,CT1	Sample analysis performed past holding time. The cooler temperature at receipt exceeded regulatory guidelines for reque
I	Semiquantitative result (out of instrument calibration range)
J	Estimated concentration; sample matrix interference.
J	Estimated value ; the analyte concentration was greater than the highest standard
J	Estimated value ; the analyte concentration was less than the LOQ.
J	The reported result is an estimated value.
J,B	Analyte detected in both the method blank and sample above the MDL.
J,CT1	Estimated value ; the analyte concentration was less than the LOQ. Cooler temperature at sample reciept exceeded regu
J,H1	Estimated value ; the analyte concentration was less than the LOQ. Sample analysis performed past holding time.
J,H1	The reported result is an estimated value. Sample was analyzed past holding time.
J,P	Estimate; columns don't agree to within 40%
J,S	Estimated concentration; analyzed by method of standard addition (MSA)
JB	The reported result is an estimated value. The reported result is also associated with a contaminated method blank.
JQ	The reported result is an estimated value and one or more quality control criteria failed. See narrative.
L	Sample reporting limits elevated due to matrix interference
L1	The associated blank spike (LCS) recovery was above the laboratory acceptance limits.
L2	The associated blank spike (LCS) recovery was below the laboratory acceptance limits.
M	Matrix effect; the concentration is an estimate due to matrix effect.
N	Nontarget analyte; the analyte is a tentativlely identified compound (TIC) by GC/MS
NA	Not applicable
ND	Not detected at or above the reporting limit (RL)
ND, B	Not detected at or above the reporting limit (RL). Analyte present in method blank.
ND, CT1	Analyte was not detected. The concentration is below the reported LOD. The cooler temperature at receipt exceeded reg
ND, L	Not detected; sample reporting limit (RL) elevated due to interference
ND, S	Not detected; analyzed by method of standard addition (MSA)
ND,H1	Not detected; Sample analysis performed past holding time.
ND,H1,CT1	Not detected; Sample analysis performed past holding time. The cooler temperature at receipt exceeded regulatory guide
NF	Not found by library search
NFL	No free liquid
NI	Non-ignitable
NR	Analyte is not required to be analyzed
NS	Not spiked
P	Concentrations >40% difference between the two GC columns
Q	One or more quality control criteria failed. See narrative.
Q,H1	One or more quality control criteria failed. Sample analyzed past holding time. See narrative.
QNS	Quantity of sample not sufficient to perform analysis
RA	Reanalysis confirms reported results
RE	Reanalysis confirms sample matrix interference
S	Analyzed by method of standard addition (MSA)
SMI	Sample matrix interference on surrogate
SP	Reported results are for spike compounds only
T5	Laboratory not licensed for this parameter
TIC	Library Search Compound
TNTC	Too numerous to count

Microbac Laboratories Inc.
List of Valid Qualifiers
July 06, 2016

Qualkey: DOD

TNTC, B	Too numerous to count. Analyte present in method blank.
TNTC,CT1	Too numerous to count. The cooler temperature at receipt exceeded regulatory guidelines for requested testing.
TNTC,H1	Too numerous to count. Sample analysis performed past holding time.
U	Analyte was not detected. The concentration is below the reported LOD.
U,CT1	Analyte was not detected. The concentration is below the reported LOD. Cooler temperature at sample reciept exceeded
U,H1	Not detected; Sample analysis performed past holding time.
UJ	Undetected; the MDL and RL are estimated due to quality control discrepancies.
UQ	Undetected; the analyte was analyzed for, but not detected.
W	Post-digestion spike for furnace AA out of control limits
X	Exceeds regulatory limit
X, S	Exceeds regulatory limit; method of standard additions (MSA)
Z	Cannot be resolved from isomer - see below

AECOM

Chain of Custody Record

COC Number:

Laboratory:	Microbac POC: Stephanie Mossburg	Project Manager:	Debra Richmann
Address:	158 Starlite Drive	Phone/Fax Number:	210-296-2000
	Marietta, OH 45750	Sampler (print):	Scott Beesinger
Phone:	1-800-373-4071		
Client:	AECOM	Signature:	Scott Beesinger
Address:	112 East Pecan Ste. 400		
	San Antonio, TX 78205		
Turn Around Time:	STANDARD	pH:	
Project Name/Location:	Longhorn		
Project Number:	60274185.0013BH 1824 MR10		

Mail to: Linda Raabe, 112 East Pecan STE. 400, San Antonio, TX 78205, 210-296-2000

Fed Ex Airbill No:

Program:

ERPIMS REQUIRED FIELDS

Site Name	Sample ID/Location ID	SBD	SED	Date	Time	Comp*	Grab	Matrix	Number of Containers	VOC	Perchlorate	Total Metals	1,4 - DIOXANE	SA CODE	Cooler ID	LOT CONTROL NUMBERS ABLOT	EBLOT	TBLOT
SITE 18/24	18CPTMW23SW-061516			6/15/16	0800		✓	W	6	✓	✓		✓					
	18CPTMW23-061516			6/15/16	0905		✓	W	6	✓	✓		✓					
	18CPTMW19SW-061516			6/15/16	1015		✓	W	6	✓	✓		✓					
	18CPTMW19-061516			6/15/16	1115		✓	W	6	✓	✓		✓					
	TRIP BLANK			6/15/16			✓	W	2	✓								

Comments: **STANDARD TAT**

	Date	Time		Date	Time	Relinquished by: (Signature)
Relinquished by: (Signature) Scott Beesinger	6/15/16	1545				
Relinquished by: (Signature)	Date	Time		Time	Remarks:	

Microbac OVD
Received: 06/16/2016 10:10
By: BRENDA GREENWALT
221000087284

Brenda Greenwalt ... Manager, Pink QA/QC Manager

*Homogenize all composite samples prior to analysis

00714660

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060927
Account: 2551
Project: 2551.096
Samples: 5
Due Date: 27-JUN-2016

Samplenum	Container ID	Products
L16060927-01	761670	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	ORG4	16-JUN-2016 15:16	CLS		
2	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Samplenum	Container ID	Products
L16060927-01	761671	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	PREP	W1	EXT	20-JUN-2016 16:36	CPD	BRG	
3	ANALYZ*	EXT	SEMI	23-JUN-2016 09:21	SCB	CPD	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	STORE	W1	A1	24-JUN-2016 08:22	BRG	BRG	

Sample extract/digestate/leachate

Samplenum	Container ID	Products
L16060927-01	761672	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	ANALYZ	W1	SEM	24-JUN-2016 12:12	JWR	BRG	
3	STORE	SEM	A1	29-JUN-2016 13:20	CLS	JWR	

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060927
Account: 2551
Project: 2551.096
Samples: 5
Due Date: 27-JUN-2016

Samplenum	Container ID	Products
L16060927-02	761673	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Samplenum	Container ID	Products
L16060927-02	761674	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	PREP	W1	EXT	20-JUN-2016 16:36	CPD	BRG	
3	ANALYZ*	EXT	SEMI	23-JUN-2016 09:21	SCB	CPD	
4	DISP	EXT	DISP	24-JUN-2016 11:59	BJO	BJO	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	STORE	W1	A1	24-JUN-2016 08:22	BRG	BRG	

Sample extract/digestate/leachate

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060927
Account: 2551
Project: 2551.096
Samples: 5
Due Date: 27-JUN-2016

Samplenum	Container ID	Products
L16060927-02	761675	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	ANALYZ	W1	SEM	24-JUN-2016 12:12	JWR	BRG	
3	STORE	SEM	A1	29-JUN-2016 13:20	CLS	JWR	

Samplenum	Container ID	Products
L16060927-03	761676	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	ORG4	16-JUN-2016 15:16	CLS		
2	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Samplenum	Container ID	Products
L16060927-03	761677	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	PREP	W1	EXT	20-JUN-2016 16:36	CPD	BRG	
3	ANALYZ*	EXT	SEMI	23-JUN-2016 09:21	SCB	CPD	
4	DISP	EXT	DISP	24-JUN-2016 11:59	BJO	BJO	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	STORE	W1	A1	24-JUN-2016 08:22	BRG	BRG	

Sample extract/digestate/leachate

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060927
Account: 2551
Project: 2551.096
Samples: 5
Due Date: 27-JUN-2016

Samplenum	Container ID	Products
L16060927-03	761678	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	ANALYZ	W1	SEM	24-JUN-2016 12:12	JWR	BRG	
3	STORE	SEM	A1	29-JUN-2016 13:20	CLS	JWR	

Samplenum	Container ID	Products
L16060927-04	761679	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Bottle: 3

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

Microbac Laboratories Inc.

Internal Chain of Custody Report

Login: L16060927
Account: 2551
Project: 2551.096
Samples: 5
Due Date: 27-JUN-2016

Samplenum	Container ID	Products
L16060927-04	761680	827-DIOXANE

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	PREP	W1	EXT	20-JUN-2016 16:36	CPD	BRG	
3	ANALYZ*	EXT	SEMI	23-JUN-2016 09:21	SCB	CPD	
4	DISP	EXT	DISP	24-JUN-2016 11:59	BJO	BJO	

Sample extract/digestate/leachate

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	STORE	W1	A1	24-JUN-2016 08:23	BRG	BRG	

Sample extract/digestate/leachate

Samplenum	Container ID	Products
L16060927-04	761681	6850

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	W1	16-JUN-2016 15:16	CLS		
2	ANALYZ	W1	SEM	24-JUN-2016 12:12	JWR	BRG	
3	STORE	SEM	A1	29-JUN-2016 13:20	CLS	JWR	

Samplenum	Container ID	Products
L16060927-05	761682	826-LOW

Bottle: 1

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

Bottle: 2

Seq.	Purpose	From	To	Date/Time	Accept	Relinquish	pH
1	LOGIN	COOLER	V1	16-JUN-2016 15:16	CLS		
2	ANALYZ	V1	ORG4	17-JUN-2016 07:41	AWE	CLS	
3	STORE	ORG4	A1	30-JUN-2016 16:12	CLS	JST	

A1 - Sample Archive (COLD)
A2 - Sample Archive (AMBIENT)
F1 - Volatiles Freezer in Login
V1 - Volatiles Refrigerator in Login
W1 - Walkin Cooler in Login

NELAP Addendum - January 4, 2016

Non-NELAP LIMS Product and Description

The following is a list of those tests that are not included in the Microbac – OVD NELAP Scope of Accreditation:

Heat of Combustion (BTU)
Total Halide by Bomb Combustion (TX)
Particle Sizing - 200 Mesh (PS200)
Specific Gravity/Density (SPGRAV)
Total Residual Chlorine (CL-TRL)
Total Volatile Solids (all forms) (TVS)
Total Coliform Bacteria (all methods)
Fecal Coliform Bacteria (all methods)
Sulfite (SO3)
Propionaldehyde (HPLC-UV)

SOLID AND HAZARDOUS CHEMICALS
Nitrogen, Ammonia by Method 350.1
Chromium, Hexavalent, Leachable by SM3500 Cr-B 2009
Phenolics, Total by Method 420.1
ASTM D3987-06

NELAP Accreditation by Laboratory SOP

NONPOTABLE WATER

OVD HPLC02/HPLC-UV

Nitroglycerin
Acetic acid
Butyric acid
Lactic acid
Propionic acid
Pyruvic acid

OVD MSS01/GC-MS

1,4-Phenylenediamine
1-Methylnaphthalene
1,4-Dioxane
Atrazine
Benzaldehyde
Biphenyl
Caprolactam
Hexamethylphosphoramide (HMPA)
Pentachlorobenzene
Pentachloroethane

NELAP Accreditation by Laboratory SOP

NONPOTABLE WATER

OVD MSV01/GC-MS

1, 1, 2-Trichloro-1,2,2-trifluoroethane
1,3-Butadiene
Cyclohexane
Cyclohexanone
Dimethyl disulfide
Dimethylsulfide
Ethyl-t-butylether (ETBE)
Isoprene
Methylacetate
Methylcyclohexane
T-amylmethylether (TAME)
Tetrahydrofuran (THF)

OVD HPLC07/HPLC-MS-MS

Hexamethylphosphoramide (XMPA-LCMS)

OVD HPLC12/HPLC/UV

Acetate
Formate

OVD RSK01/GC-FID

Acetylene
Propane

OVD K9305/ISE

Fluoroborate

SOLID AND HAZARDOUS CHEMICALS

OVD MSS0I/GC-MS

1-Methylnaphthalene
Benzaldehyde
Biphenyl
Caprolactam
Pentachloroethane

NELAP Accreditation by Laboratory SOP

SOLID AND HAZARDOUS CHEMICALS

OVD MSV0I/GC-MS

1.3-Butadiene
Cyclohexane
Cyclohexanone
Dimethyl disulfide
Dimethylsulfide
Ethyl-t-butylether (ETBE)
Isoprene
Methylacetate
Methylcyclohexane
n-Hexane
T-amylmethylether (TAME)

Microbac

Laboratory Report Number: L16061026

Kayla Teague
AECOM Technical Services, Inc.
16000 Dallas Parkway
Dallas, TX 75248

Please find enclosed the analytical results for the samples you submitted to Microbac Laboratories. Review and compilation of your report was completed by Microbac's Ohio Valley Division (OVD). If you have any questions, comments, or require further assistance regarding this report, please contact your service representative listed below.

Laboratory Contact:
Michelle Taylor – Client Services Specialist
(740) 373-4071
Michelle.Taylor@microbac.com

I certify that all test results meet all of the requirements of the DoD QSM and other applicable contract terms and conditions. Any exceptions are attached to this cover page or addressed in the method narratives presented in the report. All results for soil samples are reported on a 'dry-weight' basis unless specified otherwise. Analytical results for water and wastes are reported on a 'as received' basis unless specified otherwise. A statement of uncertainty for each analysis is available upon request. This laboratory report shall not be reproduced, except in full, without the written approval of Microbac Laboratories, DoD ELAP certification number 2936.01. The reported results are related only to the samples analyzed as received.

This report was certified on July 12 2016

David Vandenberg – Managing Director

State of Origin: TX
Accrediting Authority: Texas Commission on Environmental Quality ID:T104704252-07-TX
QAPP: DOD Ver 4.1

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Record of Sample Receipt and Inspection

Comments/Discrepancies

This is the record of the shipment conditions and the inspection records for the samples received and reported as a sample delivery group (SDG). All of the samples were inspected and observed to conform to our receipt policies, except as noted below.

There were no discrepancies.

Discrepancy	Resolution

Coolers

Cooler #	Temperature Gun	Temperature	COC #	Airbill #	Temp Required?
0019780	H	5.0		J2317164975	X

Inspection Checklist

#	Question	Result
1	Were shipping coolers sealed?	Yes
2	Were custody seals intact?	Yes
3	Were cooler temperatures in range of 0-6?	Yes
4	Was ice present?	Yes
5	Were COC's received/information complete/signed and dated?	Yes
6	Were sample containers intact and match COC?	Yes
7	Were sample labels intact and match COC?	Yes
8	Were the correct containers and volumes received?	Yes
9	Were samples received within EPA hold times?	Yes
10	All samples were checked for pH and met the standard. Exceptions are noted above under discrepancy. (water only)	Yes
11	Were pH ranges acceptable? (voa's excluded)	Yes
12	Were VOA samples free of headspace (less than 6mm)?	Yes

Microbac Laboratories ● Ohio Valley Division
158 Starlite Drive, Marietta, OH 45750 ● T: (740)373-4071 F: (740)373-4835
www.microbac.com

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Samples Received

Client ID	Laboratory ID	Date Collected	Date Received
18WW16-061616	L16061026-01	06/16/2016 07:55	06/17/2016 09:48
18WW16F-061616	L16061026-02	06/16/2016 07:55	06/17/2016 09:48
102-061616	L16061026-03	06/16/2016 08:40	06/17/2016 09:48
129-061616	L16061026-04	06/16/2016 09:05	06/17/2016 09:48
120-061616	L16061026-05	06/16/2016 09:25	06/17/2016 09:48
120F-061616	L16061026-06	06/16/2016 09:25	06/17/2016 09:48
123-061616	L16061026-07	06/16/2016 09:55	06/17/2016 09:48
TRIP BLANK	L16061026-08	06/16/2016 00:01	06/17/2016 09:48

Microbac Laboratories ● Ohio Valley Division
158 Starlite Drive, Marietta, OH 45750 ● T: (740)373-4071 F: (740)373-4835
www.microbac.com

Microbac REPORT L16061026
PREPARED FOR AECOM Technical Services, Inc.
WORK ID:

1.0 Summary Data 5
1.1 Narratives 6
1.2 Certificate of Analysis 43
2.0 Full Sample Data Package 90
2.1 Volatiles Data 91
2.1.1 Volatiles GCMS Data (8260) 92
2.1.1.1 Summary Data 93
2.1.1.2 QC Summary Data 114
2.1.1.3 Sample Data 150
2.1.1.4 Standards Data 198
2.1.1.5 Raw QC Data 253
2.2 Semivolatiles Data 354
2.2.1 Semivolatiles GC/MS Data (Dioxane) 267
2.2.1.1 Summary Data 268
2.2.1.2 QC Summary Data 271
2.2.1.3 Sample Data 301
2.2.1.4 Standards Data 311
2.2.1.5 Raw QC Data 335
2.2 General Chromatography Data 354
2.2.2 6850 LC/MS Data 355
2.2.2.1 Summary Data 356
2.2.2.2 QC Summary Data 359
2.2.2.3 Sample Data 431
2.2.2.4 Standards Data 442
2.2.2.5 Raw QC Data 503
2.3 Metals Data 510
2.3.1 Metals I C P Data 511
2.3.1.1 Summary Data 512
2.3.1.2 QC Summary Data 521
2.3.1.3 Raw Data 558
2.3.2 Metals ICP-MS Data 668
2.3.2.1 Summary Data 669
2.3.2.2 QC Summary Data 682
2.3.2.3 Raw Data 779
2.3.3 Metals CVAA Data (Mercury) 2399
2.3.3.1 Summary Data 2400
2.3.3.2 QC Summary 2407
2.3.3.3 Raw Data 2430
3.0 Attachments 2436

1.0 Summary Data

1.1 Narratives

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	8260
Prep Batch Number(s):	573776,573953	**Reviewer Name:**	Sarah Vandenberg
LRC Date:	2016-06-29 00:00:00		

Laboratory Data Package Cover Page

X	R1	Field chain-of-custody documentation;
X	R2	Sample identification cross-reference;
X	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
X	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
X	R5	Test reports/summary forms for blank samples;
X	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
X	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
X	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
X	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
X	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Sarah Vandenberg	Sarah Vandenberg		2016-06-29 20:07:04

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	8260
Prep Batch Number(s):	573776,573953	**Reviewer Name:**	Sarah Vandenberg
LRC Date:	2016-06-29 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports					
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?	X				
Were % moisture (or solids) reported for all soil and sediment samples?	X				
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?	X				
Were surrogate percent recoveries in all samples within the laboratory QC limits?	X				
Test reports/summary forms for blank samples	X				
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):					
Were all COCs included in the LCS?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	8260
Prep Batch Number(s):	573776,573953	**Reviewer Name:**	Sarah Vandenberg
LRC Date:	2016-06-29 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?	X				
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?	X				
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	8260
Prep Batch Number(s):	573776,573953	**Reviewer Name:**	Sarah Vandenberg
LRC Date:	2016-06-29 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?	X				
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?		X			1
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?			X		
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?	X				
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?	X				
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?			X		
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?			X		
Method detection limit (MDL) studies					

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	8260
Prep Batch Number(s):	573776,573953	**Reviewer Name:**	Sarah Vandenberg
LRC Date:	2016-06-29 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	8260
Prep Batch Number(s):	573776,573953	**Reviewer Name:**	Sarah Vandenberg
LRC Date:	2016-06-29 00:00:00		

below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

1)Bromomethane and 2-Hexanone exceeded the UCL in the CCV analyzed 06/22/106 om HPMS11.

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-29 00:00:00		

Laboratory Data Package Cover Page

X	R1	Field chain-of-custody documentation;
X	R2	Sample identification cross-reference;
X	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
X	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
X	R5	Test reports/summary forms for blank samples;
X	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
X	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
X	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
X	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
X	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Eric Lawson		Chemist III	2016-06-29 15:41:36

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-29 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports					
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?			X		
Were % moisture (or solids) reported for all soil and sediment samples?			X		
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?	X				
Were surrogate percent recoveries in all samples within the laboratory QC limits?	X				
Test reports/summary forms for blank samples	X				
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):					
Were all COCs included in the LCS?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-29 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?	X				
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?	X				
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-29 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?	X				
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?	X				
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?			X		
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?	X				
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?	X				
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?			X		
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?			X		
Method detection limit (MDL) studies					

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-29 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	827-DIOXANE
Prep Batch Number(s):	WG573299	**Reviewer Name:**	Eric Lawson
LRC Date:	2016-06-29 00:00:00		

below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Laboratory Data Package Cover Page

	R1	Field chain-of-custody documentation;
	R2	Sample identification cross-reference;
	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
	R5	Test reports/summary forms for blank samples;
	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Wade DeLong	Wade DeLong	Chemist I	2016-06-28 17:49:28

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports					
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?			X		
Were % moisture (or solids) reported for all soil and sediment samples?			X		
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?			X		
Were surrogate percent recoveries in all samples within the laboratory QC limits?			X		
Test reports/summary forms for blank samples	X				
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):					
Were all COCs included in the LCS?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?	X				
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?			X		
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?	X				
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?	X				
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?	X				
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?			X		
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?	X				
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?	X				
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?	X				
Method detection limit (MDL) studies					

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6850
Prep Batch Number(s):	WG574101	**Reviewer Name:**	Wade DeLong
LRC Date:	2016-06-28 00:00:00		

below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6010
Prep Batch Number(s):	WG573656	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Laboratory Data Package Cover Page

	R1	Field chain-of-custody documentation;
	R2	Sample identification cross-reference;
	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
	R5	Test reports/summary forms for blank samples;
	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Maren Beery	Maren Beery	Metals Supervisor	2016-07-11 15:22:46

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6010
Prep Batch Number(s):	WG573656	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports	X				
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				ER#1
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?	X				
Were % moisture (or solids) reported for all soil and sediment samples?	X				
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?			X		
Were surrogate percent recoveries in all samples within the laboratory QC limits?			X		
Test reports/summary forms for blank samples					
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):	X				
Were all COCs included in the LCS?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6010
Prep Batch Number(s):	WG573656	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?			X		
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?	X				
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6010
Prep Batch Number(s):	WG573656	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?					
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?	X				
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?	X				
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?			X		
Were ion abundance data within the method-required QC limits?			X		
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?			X		
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?			X		
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?	X				
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?	X				
Method detection limit (MDL) studies					

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6010
Prep Batch Number(s):	WG573656	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6010
Prep Batch Number(s):	WG573656	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

ER#1 - Client sample 02 required dilution analysis in order to obtain results for calcium, iron, magnesium and sodium within the calibration range. Client samples 04 and 06 required dilution analyses in order to obtain results for sodium within the calibration range.

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6020
Prep Batch Number(s):	WG573548	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Laboratory Data Package Cover Page

	R1	Field chain-of-custody documentation;
	R2	Sample identification cross-reference;
	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
	R5	Test reports/summary forms for blank samples;
	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Maren Beery	Maren Beery	Metals Supervisor	2016-07-11 14:48:25

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6020
Prep Batch Number(s):	WG573548	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports	X				
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				ER#1
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?	X				
Were % moisture (or solids) reported for all soil and sediment samples?	X				
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?			X		
Were surrogate percent recoveries in all samples within the laboratory QC limits?			X		
Test reports/summary forms for blank samples					
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?		X			ER#3
Laboratory control samples (LCS):					
Were all COCs included in the LCS?	X				

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6020
Prep Batch Number(s):	WG573548	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?			X		
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?	X				ER#2
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6020
Prep Batch Number(s):	WG573548	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?					
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?		X			ER#4
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?	X				
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?	X				
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?			X		
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?	X				
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?	X				
Method detection limit (MDL) studies					

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6020
Prep Batch Number(s):	WG573548	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	6020
Prep Batch Number(s):	WG573548	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

ER#1 - Client sample 02 required dilution analysis in order to obtain results for chromium, cobalt, copper, manganese, and nickel within the calibration range. Client sample 04 required dilution analysis in order to obtain a result for barium within the calibration range.

ER#2 - Due to a result that was noncompliant on the negative side upon initial analysis, vanadium for client sample 02 was reported from a dilution analysis.

ER#3 - The method blank analyzed on 29-JUN-2016 yielded a result for manganese greater than the LOD and less than the LOQ. However, all client samples had manganese concentrations that were greater than ten times that of the method blank.

ER#4 - Due to continuing calibration verification failure for cobalt on 29-JUN-2016 at 13:25, all QA/QC and client samples were reanalyzed on a later calibration for cobalt. Due to low level continuing calibration verification failure for chromium on 05-Jul-2016 at 19:20, client sample 02 was reanalyzed on a later calibration which was compliant for chromium.

ER#5 - The CCB analyzed on 29-JUN-2016 at 13:29 yielded a result for manganese greater than the LOD and less than the LOQ. However, all client samples had manganese concentrations that were greater than ten times that of the calibration blank.

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	7471
Prep Batch Number(s):	WG573569	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Laboratory Data Package Cover Page

	R1	Field chain-of-custody documentation;
	R2	Sample identification cross-reference;
	R3	Test reports (analytical data sheets) for each environmental sample that includes: (a) Items consistent with NELAC Chapter 5, (b) dilution factors, (c) preparation methods, (d) cleanup methods, and (e) a.if required for the project, tentatively identified compounds (TICs).
	R4	Surrogate recovery data including: (a) Calculated recovery (%R), and (b) the laboratory's surrogate QC limits.
	R5	Test reports/summary forms for blank samples;
	R6	Test reports/summary forms for laboratory control samples (LCSs) including: (a) LCS spiking amounts, (b) calculated %R for each analyte, and (c) the laboratory's LCS QC limits.
	R7	Test reports for project matrix spike/matrix spike duplicates (MS/MSDs) including: (a) samples associated with the MS/MSD clearly identified, (b) MS/MSD spiking compounds, (c) concentration of each MS/MSD analyte measured in the parent and spiked samples, (d) calculated %Rs and relative percent differences (RPDs), and (e) the laboratory's MS/MSD QC limits.
	R8	Laboratory analytical duplicate (if applicable) recovery and precision: (a) the amount of analyte measured in the duplicate, (b) the calculated RPD, and (c) the laboratory's QC limits for analytical duplicates.
	R9	List of method quantitation limits (MQLs) and detectability check sample results for each analyte for each method and matrix.
	R10	Other problems or anomalies.

Name (Printed)	Signature	Official Title (Printed)	Date
Maren Beery	Maren Beery	Metals Supervisor	2016-07-11 14:42:01

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	7471
Prep Batch Number(s):	WG573569	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Description	Yes	No	NA	NR	ER#
Chain-of-custody (C-O-C)					
Did samples meet the laboratory's standard conditions of sample acceptability upon receipt?	X				
Were all departures from standard conditions described in an exception report?	X				
Sample and quality control (QC) identification	X				
Are all field sample ID numbers cross-referenced to the laboratory ID numbers?	X				
Are all laboratory ID numbers cross-referenced to the corresponding QC data?	X				
Test reports	X				
Were all samples prepared and analyzed within holding times?	X				
Other than those results < MQL, were all other raw values bracketed by calibration standards?	X				
Were calculations checked by a peer or supervisor?	X				
Were all analyte identifications checked by a peer or supervisor?	X				
Were sample detection limits reported for all analytes not detected?	X				
Were all results for soil and sediment samples reported on a dry weight basis?	X				
Were % moisture (or solids) reported for all soil and sediment samples?	X				
Were bulk soils/solids samples for volatile analysis extracted with methanol per SW846 Method 5035?			X		
If required for the project, are TICs reported?			X		
Surrogate recovery data					
Were surrogates added prior to extraction?			X		
Were surrogate percent recoveries in all samples within the laboratory QC limits?			X		
Test reports/summary forms for blank samples					
Were appropriate type(s) of blanks analyzed?	X				
Were blanks analyzed at the appropriate frequency?	X				
Were method blanks taken through the entire analytical process, including preparation and, if applicable, cleanup procedures?	X				
Were blank concentrations < MQL?	X				
Laboratory control samples (LCS):	X				
Were all COCs included in the LCS?	X				

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	7471
Prep Batch Number(s):	WG573569	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Was each LCS taken through the entire analytical procedure, including prep and cleanup steps?	X				
Were LCSs analyzed at the required frequency?	X				
Were LCS (and LCSD, if applicable) %Rs within the laboratory QC limits?	X				
Does the detectability check sample data document the laboratory's capability to detect the COCs at the MDL used to calculate the SDLs?	X				
Was the LCSD RPD within QC limits?			X		
Matrix spike (MS) and matrix spike duplicate (MSD) data					
Were the project/method specified analytes included in the MS and MSD?			X		
Were MS/MSD analyzed at the appropriate frequency?			X		
Were MS (and MSD, if applicable) %Rs within the laboratory QC limits?			X		
Were MS/MSD RPDs within laboratory QC limits?			X		
Analytical duplicate data					
Were appropriate analytical duplicates analyzed for each matrix?			X		
Were analytical duplicates analyzed at the appropriate frequency?			X		
Were RPDs or relative standard deviations within the laboratory QC limits?			X		
Method quantitation limits (MQLs):					
Are the MQLs for each method analyte included in the laboratory data package?	X				
Do the MQLs correspond to the concentration of the lowest non-zero calibration standard?	X				
Are unadjusted MQLs and DCSs included in the laboratory data package?	X				
Other problems/anomalies					
Are all known problems/anomalies/special conditions noted in this LRC and ER?	X				
Was applicable and available technology used to lower the SDL to minimize the matrix interference effects on the sample results?	X				
Is the laboratory NELAC-accredited under the Texas Laboratory Accreditation Program for the analytes, matrices and methods associated with this laboratory data package?	X				
Initial calibration (ICAL)					
Were response factors and/or relative response factors for each analyte within QC limits?	X				
Were percent RSDs or correlation coefficient criteria met?	X				

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	7471
Prep Batch Number(s):	WG573569	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Was the number of standards recommended in the method used for all analytes?	X				
Were all points generated between the lowest and highest standard used to calculate the curve?					
Are ICAL data available for all instruments used?	X				
Has the initial calibration curve been verified using an appropriate second source standard?	X				
Initial and continuing calibration verification (ICCV and CCV) and continuing calibration blank (CCB):					
Was the CCV analyzed at the method-required frequency?	X				
Were percent differences for each analyte within the method-required QC limits?	X				
Was the ICAL curve verified for each analyte?	X				
Was the absolute value of the analyte concentration in the inorganic CCB < MDL?	X				
Mass spectral tuning					
Was the appropriate compound for the method used for tuning?	X				
Were ion abundance data within the method-required QC limits?	X				
Internal standards (IS)					
Were IS area counts and retention times within the method-required QC limits?	X				
Raw data (NELAC Section 5.5.10)					
Were the raw data (for example, chromatograms, spectral data) reviewed by an analyst?	X				
Were data associated with manual integrations flagged on the raw data?			X		
Dual column confirmation					
Did dual column confirmation results meet the method-required QC?			X		
Tentatively identified compounds (TICs)					
If TICs were requested, were the mass spectra and TIC data subject to appropriate checks?			X		
Interference Check Sample (ICS) results					
Were percent recoveries within method QC limits?	X				
Serial dilutions, post digestion spikes, and method of standard additions					
Were percent differences, recoveries, and the linearity within the QC limits specified in the method?	X				
Method detection limit (MDL) studies					

Microbac

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	7471
Prep Batch Number(s):	WG573569	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

Was a MDL study performed for each reported analyte?	X				
Is the MDL either adjusted or supported by the analysis of DCSs?	X				
Proficiency test reports					
Was the laboratory's performance acceptable on the applicable proficiency tests or evaluation studies?	X				
Standards documentation					
Are all standards used in the analyses NIST-traceable or obtained from other appropriate sources?	X				
Compound/analyte identification procedures					
Are the procedures for compound/analyte identification documented?	X				
Demonstration of analyst competency (DOC)					
Was DOC conducted consistent with NELAC Chapter 5?	X				
Is documentation of the analyst's competency up-to-date and on file?	X				
Verification/validation documentation for methods (NELAC Chapter 5)					
Are all the methods used to generate the data documented, verified, and validated, where applicable?	X				
Laboratory standard operating procedures (SOPs)					
Are laboratory SOPs current and on file for each method performed	X				

1. Items identified by the letter “R” must be included in the laboratory data package submitted in the TRRP-required report(s). Items identified by the letter “S” should be retained and made available upon request for the appropriate retention period;
2. O = organic analyses; I = inorganic analyses (and general chemistry, when applicable);
3. NA = Not applicable;
4. NR = Not reviewed;
5. ER# = Exception Report identification number (an Exception Report should be completed for an item if “NR” or “No” is checked).

The Exception Report for each “No” or “Not Reviewed (NR)” item in Laboratory Review Checklist and for each analyte, matrix, and method for which the laboratory does not hold NELAC accreditation under the Texas Laboratory Accreditation Program.

Release Statement: I am responsible for the release of this laboratory data package. This laboratory is NELAC accredited under the Texas Laboratory Accreditation Program for all the methods, analytes, and matrices reported in this data package except as noted in the Exception Reports. The data have been reviewed and are technically compliant with

Texas Risk Reduction Program (TRRP) Checklist

Laboratory Name:	Microbac OVD	**Laboratory Log Number:**	L16061026
Project Name:		**Method:**	7471
Prep Batch Number(s):	WG573569	**Reviewer Name:**	Maren Beery
LRC Date:	2016-07-11 00:00:00		

the requirements of the methods used, except where noted by the laboratory in the Exception Reports. By my signature below, I affirm to the best of my knowledge all problems/anomalies observed by the laboratory have been identified in the Laboratory Review Checklist, and no information affecting the quality of the data has been knowingly withheld.

Check, if applicable: [ ] This laboratory meets an exception under 30 TAC §25.6 and was last inspection by [ ] TCEQ or [ ] _____ on **(enter date of last inspection**). Any findings affecting the data in this laboratory data package are noted in the Exception Reports herein. The official signing the cover page of the report in which these data are used is responsible for releasing this data package and is by signature affirming the above release statement is true.

Exceptions Report

1.2 Certificate of Analysis

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-01	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	18WW16-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/23/2016 01:14
Collect Date:	06/16/2016 07:55	**Dilution:**	1	**File ID:**	11M12817
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	0.605	J	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	108	85	115	
1,2-Dichloroethane-d4	104	70	120	
Toluene-d8	100	85	120	
4-Bromofluorobenzene	91.3	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-01	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	18WW16-061616	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 19:40
Collect Date:	06/16/2016 07:55	**Dilution:**	1	**File ID:**	1LM.LM35503
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	5.76		0.400	0.200	0.100

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 11:42
Collect Date:	06/16/2016 07:55	**Dilution:**	1	**File ID:**	T3.062716.114220
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	0.331		0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Potassium, Total	7440-09-7	13.0		2.00	1.00	0.500
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 13:00
Collect Date:	06/16/2016 07:55	**Dilution:**	20	**File ID:**	T3.062716.130008
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Calcium, Total	7440-70-2	195		10.0	5.00	2.50
Iron, Total	7439-89-6	21.5		4.00	2.00	1.00
Magnesium, Total	7439-95-4	146		20.0	10.0	5.00
Sodium, Total	7440-23-5	587		20.0	10.0	5.00

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574572	**Analyst:**	BKT	**Run Date:**	06/29/2016 14:51
Collect Date:	06/16/2016 07:55	**Dilution:**	1	**File ID:**	NI.062916.145120
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00134	J	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00894		0.00200	0.00100	0.000500
Barium, Total	7440-39-3	0.186		0.00600	0.00300	0.00150
Cadmium, Total	7440-43-9	0.00645		0.00120	0.000600	0.000300
Lead, Total	7439-92-1	0.00100	U	0.00200	0.00100	0.000500
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Zinc, Total	7440-66-6	0.158		0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:05
Collect Date:	06/16/2016 07:55	**Dilution:**	20	**File ID:**	NI.070516.160533
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Cobalt, Total	7440-48-4	0.355		0.0400	0.0200	0.0100
Copper, Total	7440-50-8	0.828		0.0800	0.0400	0.0200
Manganese, Total	7439-96-5	3.12		0.0800	0.0400	0.0200
Vanadium, Total	7440-62-2	0.0200	U	0.0400	0.0200	0.0100

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:08
Collect Date:	06/16/2016 07:55	**Dilution:**	200	**File ID:**	NI.070516.160845
Sample Tag:	DL02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Nickel, Total	7440-02-0	14.3		1.60	0.800	0.400

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/06/2016 09:17
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/06/2016 15:27
Collect Date:	06/16/2016 07:55	**Dilution:**	200	**File ID:**	NI.070616.152752
Sample Tag:	DL03	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Chromium, Total	7440-47-3	4.62		0.800	0.400	0.200

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	18WW16F-061616	**Prep Method:**	7470A	**Prep Date:**	06/21/2016 12:47
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/24/2016 11:43
Workgroup #:	WG573634	**Analyst:**	AMA	**Run Date:**	06/24/2016 12:18
Collect Date:	06/16/2016 07:55	**Dilution:**	1	**File ID:**	M7.062416.121852
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-03	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	102-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/22/2016 23:06
Collect Date:	06/16/2016 08:40	**Dilution:**	1	**File ID:**	11M12813
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	105	85	115	
1,2-Dichloroethane-d4	98.7	70	120	
Toluene-d8	101	85	120	
4-Bromofluorobenzene	91.0	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-03	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	102-061616	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 01:12
Collect Date:	06/16/2016 08:40	**Dilution:**	100	**File ID:**	1LM.LM35479
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	107		40.0	20.0	10.0

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	129-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/23/2016 00:10
Collect Date:	06/16/2016 09:05	**Dilution:**	5	**File ID:**	11M12815
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	25.0	U	50.0	25.0	12.5
Benzene	71-43-2	1.25	U	5.00	1.25	0.625
Bromobenzene	108-86-1	1.25	U	5.00	1.25	0.625
Bromochloromethane	74-97-5	2.00	U	5.00	2.00	1.00
Bromodichloromethane	75-27-4	2.50	U	5.00	2.50	1.25
Bromoform	75-25-2	5.00	U	10.0	5.00	2.50
Bromomethane	74-83-9	5.00	Q	10.0	5.00	2.50
2-Butanone	78-93-3	25.0	U	50.0	25.0	12.5
n-Butylbenzene	104-51-8	2.50	U	5.00	2.50	1.25
sec-Butylbenzene	135-98-8	2.50	U	5.00	2.50	1.25
tert-Butylbenzene	98-06-6	2.50	U	5.00	2.50	1.25
Carbon disulfide	75-15-0	5.00	U	10.0	5.00	2.50
Carbon tetrachloride	56-23-5	2.50	U	5.00	2.50	1.25
Chlorobenzene	108-90-7	1.25	U	5.00	1.25	0.625
Chlorodibromomethane	124-48-1	2.50	U	5.00	2.50	1.25
Chloroethane	75-00-3	5.00	U	10.0	5.00	2.50
Chloroform	67-66-3	3.28	J	5.00	1.25	0.625
Chloromethane	74-87-3	5.00	U	10.0	5.00	2.50
2-Chlorotoluene	95-49-8	1.25	U	5.00	1.25	0.625
4-Chlorotoluene	106-43-4	2.50	U	5.00	2.50	1.25

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dibromo-3-chloropropane	96-12-8	10.0	U	25.0	10.0	5.00
1,2-Dibromoethane	106-93-4	2.50	U	5.00	2.50	1.25
Dibromomethane	74-95-3	2.50	U	5.00	2.50	1.25
1,2-Dichlorobenzene	95-50-1	1.25	U	5.00	1.25	0.625
1,3-Dichlorobenzene	541-73-1	2.50	U	5.00	2.50	1.25
1,4-Dichlorobenzene	106-46-7	1.25	U	5.00	1.25	0.625
Dichlorodifluoromethane	75-71-8	2.50	U	5.00	2.50	1.25
1,1-Dichloroethane	75-34-3	1.25	U	5.00	1.25	0.625
1,2-Dichloroethane	107-06-2	25.4		5.00	2.50	1.25
1,1-Dichloroethene	75-35-4	5.00	U	10.0	5.00	2.50
cis-1,2-Dichloroethene	156-59-2	5.02		5.00	2.50	1.25
trans-1,2-Dichloroethene	156-60-5	2.50	U	5.00	2.50	1.25
1,2-Dichloropropane	78-87-5	2.00	U	5.00	2.00	1.00
1,3-Dichloropropane	142-28-9	2.00	U	5.00	2.00	1.00
2,2-Dichloropropane	594-20-7	2.50	U	5.00	2.50	1.25
cis-1,3-Dichloropropene	10061-01-5	2.50	U	5.00	2.50	1.25
trans-1,3-Dichloropropene	10061-02-6	5.00	U	10.0	5.00	2.50
1,1-Dichloropropene	563-58-6	2.50	U	5.00	2.50	1.25
Ethylbenzene	100-41-4	2.50	U	5.00	2.50	1.25
2-Hexanone	591-78-6	25.0	Q	50.0	25.0	12.5
Hexachlorobutadiene	87-68-3	2.50	U	5.00	2.50	1.25
Isopropylbenzene	98-82-8	2.50	U	5.00	2.50	1.25
p-Isopropyltoluene	99-87-6	2.50	U	5.00	2.50	1.25
4-Methyl-2-pentanone	108-10-1	25.0	U	50.0	25.0	12.5
Methylene chloride	75-09-2	2.50	U	5.00	2.50	1.25
Naphthalene	91-20-3	2.00	U	5.00	2.00	1.00
n-Propylbenzene	103-65-1	1.25	U	5.00	1.25	0.625
Styrene	100-42-5	1.25	U	5.00	1.25	0.625
1,1,1,2-Tetrachloroethane	630-20-6	2.50	U	5.00	2.50	1.25
1,1,2,2-Tetrachloroethane	79-34-5	2.00	U	5.00	2.00	1.00
Tetrachloroethene	127-18-4	2.50	U	5.00	2.50	1.25
Toluene	108-88-3	2.50	U	5.00	2.50	1.25
1,2,3-Trichlorobenzene	87-61-6	1.50	U	5.00	1.50	0.750
1,2,4-Trichlorobenzene	120-82-1	2.00	U	5.00	2.00	1.00
1,1,1-Trichloroethane	71-55-6	2.50	U	5.00	2.50	1.25
1,1,2-Trichloroethane	79-00-5	2.50	U	5.00	2.50	1.25
Trichlorofluoromethane	75-69-4	2.50	U	5.00	2.50	1.25
1,2,3-Trichloropropane	96-18-4	5.00	U	10.0	5.00	2.50
1,2,4-Trimethylbenzene	95-63-6	2.50	U	5.00	2.50	1.25

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,3,5-Trimethylbenzene	108-67-8	2.50	U	5.00	2.50	1.25
Vinyl chloride	75-01-4	2.50	U	5.00	2.50	1.25
o-Xylene	95-47-6	2.50	U	5.00	2.50	1.25
m-,p-Xylene	179601-23-1	5.00	U	10.0	5.00	2.50

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	106	85	115	
1,2-Dichloroethane-d4	100	70	120	
Toluene-d8	102	85	120	
4-Bromofluorobenzene	94.7	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	129-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573953	**Analyst:**	JDS	**Run Date:**	06/23/2016 19:11
Collect Date:	06/16/2016 09:05	**Dilution:**	20	**File ID:**	11M12835
Sample Tag:	DL02	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Trichloroethene	79-01-6	1630		20.0	10.0	5.00

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	107	85	115	
1,2-Dichloroethane-d4	103	70	120	
Toluene-d8	100	85	120	
4-Bromofluorobenzene	94.7	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	129-061616	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 18:08
Collect Date:	06/16/2016 09:05	**Dilution:**	1	**File ID:**	15M18521
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	6.52		2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	115	20	129	

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	129-061616	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 01:31
Collect Date:	06/16/2016 09:05	**Dilution:**	10000	**File ID:**	1LM.LM35480
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	8950		4000	2000	1000

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	129-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 11:46
Collect Date:	06/16/2016 09:05	**Dilution:**	1	**File ID:**	T3.062716.114614
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	0.644		0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Calcium, Total	7440-70-2	13.2		0.500	0.250	0.125
Iron, Total	7439-89-6	1.18		0.200	0.100	0.0500
Magnesium, Total	7439-95-4	10.4		1.00	0.500	0.250
Potassium, Total	7440-09-7	0.716	J	2.00	1.00	0.500
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	129-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 13:03
Collect Date:	06/16/2016 09:05	**Dilution:**	20	**File ID:**	T3.062716.130358
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Sodium, Total	7440-23-5	146		20.0	10.0	5.00

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	129-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:11
Collect Date:	06/16/2016 09:05	**Dilution:**	1	**File ID:**	NI.070516.161156
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Cobalt, Total	7440-48-4	0.00182	J	0.00200	0.00100	0.000500

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	129-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574572	**Analyst:**	BKT	**Run Date:**	06/29/2016 14:54
Collect Date:	06/16/2016 09:05	**Dilution:**	1	**File ID:**	NI.062916.145431
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00110	J	0.00200	0.00100	0.000500
Cadmium, Total	7440-43-9	0.000383	J	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00415		0.00400	0.00200	0.00100
Copper, Total	7440-50-8	0.00173	J	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.000875	J	0.00200	0.00100	0.000500
Manganese, Total	7439-96-5	0.0478		0.00400	0.00200	0.00100
Nickel, Total	7440-02-0	0.00956		0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.00246		0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0250	U	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	129-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:15
Collect Date:	06/16/2016 09:05	**Dilution:**	20	**File ID:**	NI.070516.161508
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Barium, Total	7440-39-3	0.293		0.120	0.0600	0.0300

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	129-061616	**Prep Method:**	7470A	**Prep Date:**	06/21/2016 12:47
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/24/2016 11:43
Workgroup #:	WG573634	**Analyst:**	AMA	**Run Date:**	06/24/2016 12:26
Collect Date:	06/16/2016 09:05	**Dilution:**	1	**File ID:**	M7.062416.122629
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-05	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	120-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/23/2016 00:42
Collect Date:	06/16/2016 09:25	**Dilution:**	100	**File ID:**	11M12816
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	500	U	1000	500	250
Benzene	71-43-2	25.0	U	100	25.0	12.5
Bromobenzene	108-86-1	25.0	U	100	25.0	12.5
Bromochloromethane	74-97-5	40.0	U	100	40.0	20.0
Bromodichloromethane	75-27-4	50.0	U	100	50.0	25.0
Bromoform	75-25-2	100	U	200	100	50.0
Bromomethane	74-83-9	100	Q	200	100	50.0
2-Butanone	78-93-3	500	U	1000	500	250
n-Butylbenzene	104-51-8	50.0	U	100	50.0	25.0
sec-Butylbenzene	135-98-8	50.0	U	100	50.0	25.0
tert-Butylbenzene	98-06-6	50.0	U	100	50.0	25.0
Carbon disulfide	75-15-0	100	U	200	100	50.0
Carbon tetrachloride	56-23-5	50.0	U	100	50.0	25.0
Chlorobenzene	108-90-7	25.0	U	100	25.0	12.5
Chlorodibromomethane	124-48-1	50.0	U	100	50.0	25.0
Chloroethane	75-00-3	100	U	200	100	50.0
Chloroform	67-66-3	38.0	J	100	25.0	12.5
Chloromethane	74-87-3	100	U	200	100	50.0
2-Chlorotoluene	95-49-8	25.0	U	100	25.0	12.5
4-Chlorotoluene	106-43-4	50.0	U	100	50.0	25.0
1,2-Dibromo-3-chloropropane	96-12-8	200	U	500	200	100
1,2-Dibromoethane	106-93-4	50.0	U	100	50.0	25.0
Dibromomethane	74-95-3	50.0	U	100	50.0	25.0
1,2-Dichlorobenzene	95-50-1	25.0	U	100	25.0	12.5
1,3-Dichlorobenzene	541-73-1	50.0	U	100	50.0	25.0
1,4-Dichlorobenzene	106-46-7	25.0	U	100	25.0	12.5
Dichlorodifluoromethane	75-71-8	50.0	U	100	50.0	25.0
1,1-Dichloroethane	75-34-3	25.0	U	100	25.0	12.5
1,2-Dichloroethane	107-06-2	32.1	J	100	50.0	25.0
1,1-Dichloroethene	75-35-4	134	J	200	100	50.0
cis-1,2-Dichloroethene	156-59-2	1510		100	50.0	25.0

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	50.0	U	100	50.0	25.0
1,2-Dichloropropane	78-87-5	40.0	U	100	40.0	20.0
1,3-Dichloropropane	142-28-9	40.0	U	100	40.0	20.0
2,2-Dichloropropane	594-20-7	50.0	U	100	50.0	25.0
cis-1,3-Dichloropropene	10061-01-5	50.0	U	100	50.0	25.0
trans-1,3-Dichloropropene	10061-02-6	100	U	200	100	50.0
1,1-Dichloropropene	563-58-6	50.0	U	100	50.0	25.0
Ethylbenzene	100-41-4	50.0	U	100	50.0	25.0
2-Hexanone	591-78-6	500	Q	1000	500	250
Hexachlorobutadiene	87-68-3	50.0	U	100	50.0	25.0
Isopropylbenzene	98-82-8	50.0	U	100	50.0	25.0
p-Isopropyltoluene	99-87-6	50.0	U	100	50.0	25.0
4-Methyl-2-pentanone	108-10-1	500	U	1000	500	250
Methylene chloride	75-09-2	50.0	U	100	50.0	25.0
Naphthalene	91-20-3	40.0	U	100	40.0	20.0
n-Propylbenzene	103-65-1	25.0	U	100	25.0	12.5
Styrene	100-42-5	25.0	U	100	25.0	12.5
1,1,1,2-Tetrachloroethane	630-20-6	50.0	U	100	50.0	25.0
1,1,2,2-Tetrachloroethane	79-34-5	40.0	U	100	40.0	20.0
Tetrachloroethene	127-18-4	50.0	U	100	50.0	25.0
Toluene	108-88-3	50.0	U	100	50.0	25.0
1,2,3-Trichlorobenzene	87-61-6	30.0	U	100	30.0	15.0
1,2,4-Trichlorobenzene	120-82-1	40.0	U	100	40.0	20.0
1,1,1-Trichloroethane	71-55-6	50.0	U	100	50.0	25.0
1,1,2-Trichloroethane	79-00-5	50.0	U	100	50.0	25.0
Trichloroethene	79-01-6	17100		100	50.0	25.0
Trichlorofluoromethane	75-69-4	50.0	U	100	50.0	25.0
1,2,3-Trichloropropane	96-18-4	100	U	200	100	50.0
1,2,4-Trimethylbenzene	95-63-6	50.0	U	100	50.0	25.0
1,3,5-Trimethylbenzene	108-67-8	50.0	U	100	50.0	25.0
Vinyl chloride	75-01-4	88.7	J	100	50.0	25.0
o-Xylene	95-47-6	50.0	U	100	50.0	25.0
m-,p-Xylene	179601-23-1	100	U	200	100	50.0

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	106	85	115	
1,2-Dichloroethane-d4	103	70	120	
Toluene-d8	102	85	120	
4-Bromofluorobenzene	94.1	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-05	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	120-061616	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/27/2016 15:30
Collect Date:	06/16/2016 09:25	**Dilution:**	40	**File ID:**	15M18600
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	220		80.0	40.0	20.0

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	0.000	20	129	*

Certificate of Analysis

Sample #:	L16061026-05	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	120-061616	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 19:59
Collect Date:	06/16/2016 09:25	**Dilution:**	10000	**File ID:**	1LM.LM35504
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	32700		4000	2000	1000

Certificate of Analysis

Sample #:	L16061026-06	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	120F-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 11:50
Collect Date:	06/16/2016 09:25	**Dilution:**	1	**File ID:**	T3.062716.115009
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	1.53		0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Calcium, Total	7440-70-2	6.89		0.500	0.250	0.125
Iron, Total	7439-89-6	2.04		0.200	0.100	0.0500
Magnesium, Total	7439-95-4	7.27		1.00	0.500	0.250
Potassium, Total	7440-09-7	1.55	J	2.00	1.00	0.500
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-06	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	120F-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 13:15
Collect Date:	06/16/2016 09:25	**Dilution:**	20	**File ID:**	T3.062716.131528
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Sodium, Total	7440-23-5	303		20.0	10.0	5.00

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-06	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	120F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574572	**Analyst:**	BKT	**Run Date:**	06/29/2016 14:57
Collect Date:	06/16/2016 09:25	**Dilution:**	1	**File ID:**	NI.062916.145742
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00356		0.00200	0.00100	0.000500
Barium, Total	7440-39-3	0.0452		0.00600	0.00300	0.00150
Cadmium, Total	7440-43-9	0.000600	U	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.0120		0.00400	0.00200	0.00100
Copper, Total	7440-50-8	0.00476		0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00162	J	0.00200	0.00100	0.000500
Manganese, Total	7439-96-5	0.186		0.00400	0.00200	0.00100
Nickel, Total	7440-02-0	0.0755		0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.0105		0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0659		0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-06	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	120F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:30
Collect Date:	06/16/2016 09:25	**Dilution:**	1	**File ID:**	NI.070516.163034
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Cobalt, Total	7440-48-4	0.0151		0.00200	0.00100	0.000500

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-06	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	120F-061616	**Prep Method:**	7470A	**Prep Date:**	06/21/2016 12:47
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/24/2016 11:43
Workgroup #:	WG573634	**Analyst:**	AMA	**Run Date:**	06/24/2016 12:31
Collect Date:	06/16/2016 09:25	**Dilution:**	1	**File ID:**	M7.062416.123133
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	123-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/22/2016 23:38
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	11M12814
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	1.20		1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	14.4		1.00	0.500	0.250

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	80.1		1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	106	85	115	
1,2-Dichloroethane-d4	99.4	70	120	
Toluene-d8	100	85	120	
4-Bromofluorobenzene	93.5	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	123-061616	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 18:54
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	15M18523
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	0.587	J	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	50.8	20	129	

J	Estimated value ; the analyte concentration was less than the LOQ.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	123-061616	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 20:18
Collect Date:	06/16/2016 09:55	**Dilution:**	50	**File ID:**	1LM.LM35505
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	105		20.0	10.0	5.00

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	123-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 11:54
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	T3.062716.115400
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	1.42		0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Calcium, Total	7440-70-2	8.16		0.500	0.250	0.125
Iron, Total	7439-89-6	1.66		0.200	0.100	0.0500
Magnesium, Total	7439-95-4	4.83		1.00	0.500	0.250
Potassium, Total	7440-09-7	0.713	J	2.00	1.00	0.500
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500
Sodium, Total	7440-23-5	26.0		1.00	0.500	0.250

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	123-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574572	**Analyst:**	BKT	**Run Date:**	06/29/2016 15:00
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	NI.062916.150054
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.000767	J	0.00200	0.00100	0.000500
Barium, Total	7440-39-3	0.136		0.00600	0.00300	0.00150
Cadmium, Total	7440-43-9	0.000600	U	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00177	J	0.00400	0.00200	0.00100
Copper, Total	7440-50-8	0.00142	J	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00104	J	0.00200	0.00100	0.000500
Manganese, Total	7439-96-5	0.0161		0.00400	0.00200	0.00100
Nickel, Total	7440-02-0	0.00319	J	0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.00241		0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0250	U	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	123-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:34
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	NI.070516.163440
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Cobalt, Total	7440-48-4	0.00250		0.00200	0.00100	0.000500

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	CVAA1
Client ID:	123-061616	**Prep Method:**	7470A	**Prep Date:**	06/21/2016 12:47
Matrix:	Water	**Analytical Method:**	7470A	**Cal Date:**	06/24/2016 11:43
Workgroup #:	WG573634	**Analyst:**	AMA	**Run Date:**	06/24/2016 12:34
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	M7.062416.123404
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Mercury	7439-97-6	0.000200	U	0.000400	0.000200	0.000100

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-08	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	TRIP BLANK	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/22/2016 20:27
Collect Date:	06/16/2016 00:01	**Dilution:**	1	**File ID:**	11M12808
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	103	85	115	
1,2-Dichloroethane-d4	92.3	70	120	
Toluene-d8	101	85	120	
4-Bromofluorobenzene	93.0	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

2.0 Full Sample Data Package

2.1 Volatiles Data

2.1.1 Volatiles GCMS Data (8260)

2.1.1.1 Summary Data

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-01	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	18WW16-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/23/2016 01:14
Collect Date:	06/16/2016 07:55	**Dilution:**	1	**File ID:**	11M12817
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	0.605	J	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	108	85	115	
1,2-Dichloroethane-d4	104	70	120	
Toluene-d8	100	85	120	
4-Bromofluorobenzene	91.3	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-03	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	102-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/22/2016 23:06
Collect Date:	06/16/2016 08:40	**Dilution:**	1	**File ID:**	11M12813
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	105	85	115	
1,2-Dichloroethane-d4	98.7	70	120	
Toluene-d8	101	85	120	
4-Bromofluorobenzene	91.0	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	129-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/23/2016 00:10
Collect Date:	06/16/2016 09:05	**Dilution:**	5	**File ID:**	11M12815
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	25.0	U	50.0	25.0	12.5
Benzene	71-43-2	1.25	U	5.00	1.25	0.625
Bromobenzene	108-86-1	1.25	U	5.00	1.25	0.625
Bromochloromethane	74-97-5	2.00	U	5.00	2.00	1.00
Bromodichloromethane	75-27-4	2.50	U	5.00	2.50	1.25
Bromoform	75-25-2	5.00	U	10.0	5.00	2.50
Bromomethane	74-83-9	5.00	Q	10.0	5.00	2.50
2-Butanone	78-93-3	25.0	U	50.0	25.0	12.5
n-Butylbenzene	104-51-8	2.50	U	5.00	2.50	1.25
sec-Butylbenzene	135-98-8	2.50	U	5.00	2.50	1.25
tert-Butylbenzene	98-06-6	2.50	U	5.00	2.50	1.25
Carbon disulfide	75-15-0	5.00	U	10.0	5.00	2.50
Carbon tetrachloride	56-23-5	2.50	U	5.00	2.50	1.25
Chlorobenzene	108-90-7	1.25	U	5.00	1.25	0.625
Chlorodibromomethane	124-48-1	2.50	U	5.00	2.50	1.25
Chloroethane	75-00-3	5.00	U	10.0	5.00	2.50
Chloroform	67-66-3	3.28	J	5.00	1.25	0.625
Chloromethane	74-87-3	5.00	U	10.0	5.00	2.50
2-Chlorotoluene	95-49-8	1.25	U	5.00	1.25	0.625
4-Chlorotoluene	106-43-4	2.50	U	5.00	2.50	1.25
1,2-Dibromo-3-chloropropane	96-12-8	10.0	U	25.0	10.0	5.00
1,2-Dibromoethane	106-93-4	2.50	U	5.00	2.50	1.25
Dibromomethane	74-95-3	2.50	U	5.00	2.50	1.25
1,2-Dichlorobenzene	95-50-1	1.25	U	5.00	1.25	0.625
1,3-Dichlorobenzene	541-73-1	2.50	U	5.00	2.50	1.25
1,4-Dichlorobenzene	106-46-7	1.25	U	5.00	1.25	0.625
Dichlorodifluoromethane	75-71-8	2.50	U	5.00	2.50	1.25
1,1-Dichloroethane	75-34-3	1.25	U	5.00	1.25	0.625
1,2-Dichloroethane	107-06-2	25.4		5.00	2.50	1.25
1,1-Dichloroethene	75-35-4	5.00	U	10.0	5.00	2.50
cis-1,2-Dichloroethene	156-59-2	5.02		5.00	2.50	1.25

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	2.50	U	5.00	2.50	1.25
1,2-Dichloropropane	78-87-5	2.00	U	5.00	2.00	1.00
1,3-Dichloropropane	142-28-9	2.00	U	5.00	2.00	1.00
2,2-Dichloropropane	594-20-7	2.50	U	5.00	2.50	1.25
cis-1,3-Dichloropropene	10061-01-5	2.50	U	5.00	2.50	1.25
trans-1,3-Dichloropropene	10061-02-6	5.00	U	10.0	5.00	2.50
1,1-Dichloropropene	563-58-6	2.50	U	5.00	2.50	1.25
Ethylbenzene	100-41-4	2.50	U	5.00	2.50	1.25
2-Hexanone	591-78-6	25.0	Q	50.0	25.0	12.5
Hexachlorobutadiene	87-68-3	2.50	U	5.00	2.50	1.25
Isopropylbenzene	98-82-8	2.50	U	5.00	2.50	1.25
p-Isopropyltoluene	99-87-6	2.50	U	5.00	2.50	1.25
4-Methyl-2-pentanone	108-10-1	25.0	U	50.0	25.0	12.5
Methylene chloride	75-09-2	2.50	U	5.00	2.50	1.25
Naphthalene	91-20-3	2.00	U	5.00	2.00	1.00
n-Propylbenzene	103-65-1	1.25	U	5.00	1.25	0.625
Styrene	100-42-5	1.25	U	5.00	1.25	0.625
1,1,1,2-Tetrachloroethane	630-20-6	2.50	U	5.00	2.50	1.25
1,1,2,2-Tetrachloroethane	79-34-5	2.00	U	5.00	2.00	1.00
Tetrachloroethene	127-18-4	2.50	U	5.00	2.50	1.25
Toluene	108-88-3	2.50	U	5.00	2.50	1.25
1,2,3-Trichlorobenzene	87-61-6	1.50	U	5.00	1.50	0.750
1,2,4-Trichlorobenzene	120-82-1	2.00	U	5.00	2.00	1.00
1,1,1-Trichloroethane	71-55-6	2.50	U	5.00	2.50	1.25
1,1,2-Trichloroethane	79-00-5	2.50	U	5.00	2.50	1.25
Trichlorofluoromethane	75-69-4	2.50	U	5.00	2.50	1.25
1,2,3-Trichloropropane	96-18-4	5.00	U	10.0	5.00	2.50
1,2,4-Trimethylbenzene	95-63-6	2.50	U	5.00	2.50	1.25
1,3,5-Trimethylbenzene	108-67-8	2.50	U	5.00	2.50	1.25
Vinyl chloride	75-01-4	2.50	U	5.00	2.50	1.25
o-Xylene	95-47-6	2.50	U	5.00	2.50	1.25
m-,p-Xylene	179601-23-1	5.00	U	10.0	5.00	2.50

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	106	85	115	
1,2-Dichloroethane-d4	100	70	120	
Toluene-d8	102	85	120	
4-Bromofluorobenzene	94.7	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	129-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573953	**Analyst:**	JDS	**Run Date:**	06/23/2016 19:11
Collect Date:	06/16/2016 09:05	**Dilution:**	20	**File ID:**	11M12835
Sample Tag:	DL02	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Trichloroethene	79-01-6	1630		20.0	10.0	5.00

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	107	85	115	
1,2-Dichloroethane-d4	103	70	120	
Toluene-d8	100	85	120	
4-Bromofluorobenzene	94.7	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.
Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-05	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	120-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/23/2016 00:42
Collect Date:	06/16/2016 09:25	**Dilution:**	100	**File ID:**	11M12816
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	500	U	1000	500	250
Benzene	71-43-2	25.0	U	100	25.0	12.5
Bromobenzene	108-86-1	25.0	U	100	25.0	12.5
Bromochloromethane	74-97-5	40.0	U	100	40.0	20.0
Bromodichloromethane	75-27-4	50.0	U	100	50.0	25.0
Bromoform	75-25-2	100	U	200	100	50.0
Bromomethane	74-83-9	100	Q	200	100	50.0
2-Butanone	78-93-3	500	U	1000	500	250
n-Butylbenzene	104-51-8	50.0	U	100	50.0	25.0
sec-Butylbenzene	135-98-8	50.0	U	100	50.0	25.0
tert-Butylbenzene	98-06-6	50.0	U	100	50.0	25.0
Carbon disulfide	75-15-0	100	U	200	100	50.0
Carbon tetrachloride	56-23-5	50.0	U	100	50.0	25.0
Chlorobenzene	108-90-7	25.0	U	100	25.0	12.5
Chlorodibromomethane	124-48-1	50.0	U	100	50.0	25.0
Chloroethane	75-00-3	100	U	200	100	50.0
Chloroform	67-66-3	38.0	J	100	25.0	12.5
Chloromethane	74-87-3	100	U	200	100	50.0
2-Chlorotoluene	95-49-8	25.0	U	100	25.0	12.5
4-Chlorotoluene	106-43-4	50.0	U	100	50.0	25.0
1,2-Dibromo-3-chloropropane	96-12-8	200	U	500	200	100
1,2-Dibromoethane	106-93-4	50.0	U	100	50.0	25.0
Dibromomethane	74-95-3	50.0	U	100	50.0	25.0
1,2-Dichlorobenzene	95-50-1	25.0	U	100	25.0	12.5
1,3-Dichlorobenzene	541-73-1	50.0	U	100	50.0	25.0
1,4-Dichlorobenzene	106-46-7	25.0	U	100	25.0	12.5
Dichlorodifluoromethane	75-71-8	50.0	U	100	50.0	25.0
1,1-Dichloroethane	75-34-3	25.0	U	100	25.0	12.5
1,2-Dichloroethane	107-06-2	32.1	J	100	50.0	25.0
1,1-Dichloroethene	75-35-4	134	J	200	100	50.0
cis-1,2-Dichloroethene	156-59-2	1510		100	50.0	25.0

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	50.0	U	100	50.0	25.0
1,2-Dichloropropane	78-87-5	40.0	U	100	40.0	20.0
1,3-Dichloropropane	142-28-9	40.0	U	100	40.0	20.0
2,2-Dichloropropane	594-20-7	50.0	U	100	50.0	25.0
cis-1,3-Dichloropropene	10061-01-5	50.0	U	100	50.0	25.0
trans-1,3-Dichloropropene	10061-02-6	100	U	200	100	50.0
1,1-Dichloropropene	563-58-6	50.0	U	100	50.0	25.0
Ethylbenzene	100-41-4	50.0	U	100	50.0	25.0
2-Hexanone	591-78-6	500	Q	1000	500	250
Hexachlorobutadiene	87-68-3	50.0	U	100	50.0	25.0
Isopropylbenzene	98-82-8	50.0	U	100	50.0	25.0
p-Isopropyltoluene	99-87-6	50.0	U	100	50.0	25.0
4-Methyl-2-pentanone	108-10-1	500	U	1000	500	250
Methylene chloride	75-09-2	50.0	U	100	50.0	25.0
Naphthalene	91-20-3	40.0	U	100	40.0	20.0
n-Propylbenzene	103-65-1	25.0	U	100	25.0	12.5
Styrene	100-42-5	25.0	U	100	25.0	12.5
1,1,1,2-Tetrachloroethane	630-20-6	50.0	U	100	50.0	25.0
1,1,2,2-Tetrachloroethane	79-34-5	40.0	U	100	40.0	20.0
Tetrachloroethene	127-18-4	50.0	U	100	50.0	25.0
Toluene	108-88-3	50.0	U	100	50.0	25.0
1,2,3-Trichlorobenzene	87-61-6	30.0	U	100	30.0	15.0
1,2,4-Trichlorobenzene	120-82-1	40.0	U	100	40.0	20.0
1,1,1-Trichloroethane	71-55-6	50.0	U	100	50.0	25.0
1,1,2-Trichloroethane	79-00-5	50.0	U	100	50.0	25.0
Trichloroethene	79-01-6	17100		100	50.0	25.0
Trichlorofluoromethane	75-69-4	50.0	U	100	50.0	25.0
1,2,3-Trichloropropane	96-18-4	100	U	200	100	50.0
1,2,4-Trimethylbenzene	95-63-6	50.0	U	100	50.0	25.0
1,3,5-Trimethylbenzene	108-67-8	50.0	U	100	50.0	25.0
Vinyl chloride	75-01-4	88.7	J	100	50.0	25.0
o-Xylene	95-47-6	50.0	U	100	50.0	25.0
m-,p-Xylene	179601-23-1	100	U	200	100	50.0

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	106	85	115	
1,2-Dichloroethane-d4	103	70	120	
Toluene-d8	102	85	120	
4-Bromofluorobenzene	94.1	75	120	

J	Estimated value ; the analyte concentration was less than the LOQ.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Q	One or more quality control criteria failed. See narrative.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	123-061616	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/22/2016 23:38
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	11M12814
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	1.20		1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	14.4		1.00	0.500	0.250

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	80.1		1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	106	85	115	
1,2-Dichloroethane-d4	99.4	70	120	
Toluene-d8	100	85	120	
4-Bromofluorobenzene	93.5	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-08	**PrePrep Method:**	N/A	**Instrument:**	HPMS11
Client ID:	TRIP BLANK	**Prep Method:**	5030B/5030C/5035A	**Prep Date:**	N/A
Matrix:	Water	**Analytical Method:**	8260B	**Cal Date:**	05/13/2016 19:26
Workgroup #:	WG573776	**Analyst:**	JDS	**Run Date:**	06/22/2016 20:27
Collect Date:	06/16/2016 00:01	**Dilution:**	1	**File ID:**	11M12808
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Acetone	67-64-1	5.00	U	10.0	5.00	2.50
Benzene	71-43-2	0.250	U	1.00	0.250	0.125
Bromobenzene	108-86-1	0.250	U	1.00	0.250	0.125
Bromochloromethane	74-97-5	0.400	U	1.00	0.400	0.200
Bromodichloromethane	75-27-4	0.500	U	1.00	0.500	0.250
Bromoform	75-25-2	1.00	U	2.00	1.00	0.500
Bromomethane	74-83-9	1.00	Q	2.00	1.00	0.500
2-Butanone	78-93-3	5.00	U	10.0	5.00	2.50
n-Butylbenzene	104-51-8	0.500	U	1.00	0.500	0.250
sec-Butylbenzene	135-98-8	0.500	U	1.00	0.500	0.250
tert-Butylbenzene	98-06-6	0.500	U	1.00	0.500	0.250
Carbon disulfide	75-15-0	1.00	U	2.00	1.00	0.500
Carbon tetrachloride	56-23-5	0.500	U	1.00	0.500	0.250
Chlorobenzene	108-90-7	0.250	U	1.00	0.250	0.125
Chlorodibromomethane	124-48-1	0.500	U	1.00	0.500	0.250
Chloroethane	75-00-3	1.00	U	2.00	1.00	0.500
Chloroform	67-66-3	0.250	U	1.00	0.250	0.125
Chloromethane	74-87-3	1.00	U	2.00	1.00	0.500
2-Chlorotoluene	95-49-8	0.250	U	1.00	0.250	0.125
4-Chlorotoluene	106-43-4	0.500	U	1.00	0.500	0.250
1,2-Dibromo-3-chloropropane	96-12-8	2.00	U	5.00	2.00	1.00
1,2-Dibromoethane	106-93-4	0.500	U	1.00	0.500	0.250
Dibromomethane	74-95-3	0.500	U	1.00	0.500	0.250
1,2-Dichlorobenzene	95-50-1	0.250	U	1.00	0.250	0.125
1,3-Dichlorobenzene	541-73-1	0.500	U	1.00	0.500	0.250
1,4-Dichlorobenzene	106-46-7	0.250	U	1.00	0.250	0.125
Dichlorodifluoromethane	75-71-8	0.500	U	1.00	0.500	0.250
1,1-Dichloroethane	75-34-3	0.250	U	1.00	0.250	0.125
1,2-Dichloroethane	107-06-2	0.500	U	1.00	0.500	0.250
1,1-Dichloroethene	75-35-4	1.00	U	2.00	1.00	0.500
cis-1,2-Dichloroethene	156-59-2	0.500	U	1.00	0.500	0.250

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
trans-1,2-Dichloroethene	156-60-5	0.500	U	1.00	0.500	0.250
1,2-Dichloropropane	78-87-5	0.400	U	1.00	0.400	0.200
1,3-Dichloropropane	142-28-9	0.400	U	1.00	0.400	0.200
2,2-Dichloropropane	594-20-7	0.500	U	1.00	0.500	0.250
cis-1,3-Dichloropropene	10061-01-5	0.500	U	1.00	0.500	0.250
trans-1,3-Dichloropropene	10061-02-6	1.00	U	2.00	1.00	0.500
1,1-Dichloropropene	563-58-6	0.500	U	1.00	0.500	0.250
Ethylbenzene	100-41-4	0.500	U	1.00	0.500	0.250
2-Hexanone	591-78-6	5.00	Q	10.0	5.00	2.50
Hexachlorobutadiene	87-68-3	0.500	U	1.00	0.500	0.250
Isopropylbenzene	98-82-8	0.500	U	1.00	0.500	0.250
p-Isopropyltoluene	99-87-6	0.500	U	1.00	0.500	0.250
4-Methyl-2-pentanone	108-10-1	5.00	U	10.0	5.00	2.50
Methylene chloride	75-09-2	0.500	U	1.00	0.500	0.250
Naphthalene	91-20-3	0.400	U	1.00	0.400	0.200
n-Propylbenzene	103-65-1	0.250	U	1.00	0.250	0.125
Styrene	100-42-5	0.250	U	1.00	0.250	0.125
1,1,1,2-Tetrachloroethane	630-20-6	0.500	U	1.00	0.500	0.250
1,1,2,2-Tetrachloroethane	79-34-5	0.400	U	1.00	0.400	0.200
Tetrachloroethene	127-18-4	0.500	U	1.00	0.500	0.250
Toluene	108-88-3	0.500	U	1.00	0.500	0.250
1,2,3-Trichlorobenzene	87-61-6	0.300	U	1.00	0.300	0.150
1,2,4-Trichlorobenzene	120-82-1	0.400	U	1.00	0.400	0.200
1,1,1-Trichloroethane	71-55-6	0.500	U	1.00	0.500	0.250
1,1,2-Trichloroethane	79-00-5	0.500	U	1.00	0.500	0.250
Trichloroethene	79-01-6	0.500	U	1.00	0.500	0.250
Trichlorofluoromethane	75-69-4	0.500	U	1.00	0.500	0.250
1,2,3-Trichloropropane	96-18-4	1.00	U	2.00	1.00	0.500
1,2,4-Trimethylbenzene	95-63-6	0.500	U	1.00	0.500	0.250
1,3,5-Trimethylbenzene	108-67-8	0.500	U	1.00	0.500	0.250
Vinyl chloride	75-01-4	0.500	U	1.00	0.500	0.250
o-Xylene	95-47-6	0.500	U	1.00	0.500	0.250
m-,p-Xylene	179601-23-1	1.00	U	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
Dibromofluoromethane	103	85	115	
1,2-Dichloroethane-d4	92.3	70	120	
Toluene-d8	101	85	120	
4-Bromofluorobenzene	93.0	75	120	

Q	One or more quality control criteria failed. See narrative.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.1.1.2 QC Summary Data

Example 8260 Calculations

1.0 Calculating the Response Factor (RF) from the initial calibration (ICAL) data:

RF = [(Ax) (Cis)] / [(Ais) (Cx)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured:	3399156
Cis = Concentration of the specific internal standard (ug/mL)	25
Ais = Area of the characteristic ion of the specific internal standard	846471
Cx = Concentration of the compound in the standard being measured (ug/mL)	100
RF = Calculated Response Factor	**1.0039**

2.0 Calculating the concentration (C) of a compound in water using the average RF: *

Cx = [(Ax) (Cis) (Vn)(D)] / [(Ais) (RF) (Vs)]

where:	**Example**
Ax = Area of the characteristic ion for the compound being measured	**3122498**
Cis = Concentration of the specific internal standard (ug/L)	**25**
D = Dilution factor for sample as a multiplier (10x = 10)	**1**
Ais = Area of the characteristic ion of the specific internal standard	**611048**
RF = Average RF from the ICAL	**1.004**
Vs = Purge volume of sample (mL)	**10**
Vn = Nominal purge volume of sample (mL) (10.0 mL)	**10**
Cx = Concentration of the compound in the sample being measured (ug/L)	**127.2428**

3.0 Calculating the concentration (C) of a compound in soil using the average RF: *

Cx = [(Ax) (Cis) (Wn)(D)] / [(Ais) (RF) (Ws)]

where:	**Example**
Ax = Area of the characteristic ion for the compound being measured	**3122498**
Cis = Concentration of the specific internal standard (ug/L)	**25**
D = Dilution factor for sample as a multiplier (10x = 10)	**1**
Ais = Area of the characteristic ion of the specific internal standard	**611048**
RF = Average RF from the ICAL	**1.004**
Ws = Weight of sample purged (g)	**5**
Wn = Nominal purge weight (g) (5.0 g)	**5**
Cx = Concentration of the compound in the sample being measured (ug/L)	**127.2428**

Dry weight correction:

Percent solids (PCT_S)	**50**
Cd = (Cx) (100)/PCT_S	**254.4856**

* Concentrations appearing on the instrument quantitation reports are on-column results and do not take into account initial volume, final volume, and the dilution factor.

4.0 Concentration from Linear Regression

Step 1: Retrieve Curve Data From Plot, y = mx + b

y = response ratio = response of analyte / response of IS = Ax/Ais

x = amount ratio = concentration analyte/concentration internal standard = Cx / Cis

m = slope from curve = 0.213

b = intercept from curve = - 0.00642

Step 2: Calculate y from Quantitation Report

y = 86550/593147 = 0.1459

Step 3: Solve for x

x = (y - b)/m = [(0.1459 -(- 0.00642)]/ 0.213 = 0.7152

Step 4: Solve for analyte concentration Cx

Cx = Cis (x) = (25.0)(0.7152) = 17.88

Example Spreadsheet Calculation:

Slope from curve, m:	**0.213**
Intercept from curve, b:	**-0.00642**
Area of analyte, Ax:	**86550**
Area of Internal Standard , Ais:	**593147**
Concentration of IS, Cis	**25.00**
Response Ratio:	**0.145917**
Amount Ratio:	**0.715195**
Concentration:	**17.87988**
Units of Internal Standard:	**ug/L**

5.0 Concentration from Quadratic Regression

Step 1 - Retrieve Curve Data from Plot, y = Ax^2 + Bx + C
Where:
Ax^2 + Bx + (C - y) = 0
A, B, C = constants from the ICAL quadratic regression
y = Response ratio = Area of analyte/Area of internal standard (IS)
x = Amount ratio = Concentration of analyte/concentration of IS

Step 2: Calculate y from Quantitation Report

y = Ax/Ais

Step 3: Solve for x using the quadratic formula
Ax^2 + Bx + C - y = 0

$$x = \frac{b \pm \sqrt{\left(b^2 - 4a(c - y)\right)}}{2a}$$ (Two possible solutions)

Step 4: Solve for analyte concentration Cx

Cx = (Cis)(Amount ratio)

Example Spreadsheet Calculation:

Value of A from plot:	**-0.00629**	
Value of B from plot:	**0.511**	
Value of C from plot:	**-0.0276**	
Area of unknown from quantitation report:	**293821**	
Area of IS from quantiation report:	**784848**	
Response ratio, y:	**0.374367**	
C - y:	**-0.40197**	
Root 1 - Computed amount ratio , X1:	**80.44567**	
Root 2 - Computed amount ratio , X2:	**0.794396**	use this solution
Concentration of IS, Cis:	**25.00**	
Concentration of analyte, Cx:	**19.86**	ug/L

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS11 Dataset: 051316
Analyst1: JDS Analyst2: NA
Method: 8260B SOP: MSV01 / OVAP MSV01 Rev: 23 / 0
Method: 5030B/5030C/5035A SOP: PAT01 / OVAP PAT01 Rev: 18 / 0
Method: 624 SOP: MSV10 Rev: 14
Maintenance Log ID:

Internal Standard: STD76110 Surrogate Standard: STD75929
CCV: STD76127 LCS: STD76109 MS/MSD: NA

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG568769

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
11M11835	RINSE	NA	1	1		05/13/16 11:59
11M11836	WG568769-01 50ng BFB STD 8260	NA	1	1	STD76034	05/13/16 14:15
11M11837	WG568769-02 0.3ug/L ICAL STD 8260	NA	1	1	STD76127	05/13/16 14:40
11M11838	WG568769-03 0.4ug/L ICAL STD 8260	NA	1	1	STD76127	05/13/16 15:12
11M11839	WG568769-04 1.0ug/L ICAL STD 8260	NA	1	1	STD76127	05/13/16 15:43
11M11840	WG568769-05 2.0ug/L ICAL STD 8260	NA	1	1	STD76127	05/13/16 16:15
11M11841	WG568769-06 5.0ug/L ICAL STD 8260	NA	1	1	STD76127	05/13/16 16:47
11M11842	WG568769-07 20ug/L ICAL STD 8260	NA	1	1	STD76127	05/13/16 17:19
11M11843	WG568769-08 50ug/L ICAL STD 8260	NA	1	1	STD76127	05/13/16 17:51
11M11844	WG568769-09 100ug/L ICAL STD 8260	NA	1	1	STD76127	05/13/16 18:22
11M11845	WG568769-10 200ug/L ICAL STD 8260	NA	1	1	STD76127	05/13/16 18:54
11M11846	WG568769-11 300ug/L ICAL STD 8260	NA	1	1	STD76127	05/13/16 19:26
11M11847	RINSE	NA	1	1		05/13/16 19:58
11M11848	WG568769-12 50ug/L ALT SRC STD 8260	NA	1	1	STD76109	05/13/16 20:30
11M11849	RINSE	NA	1	1		05/13/16 21:02
11M11850	RINSE	NA	1	1		05/13/16 21:34

Approved: May 23, 2016

Page: 1

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS11 Dataset: 062216
Analyst1: JDS Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 24
Method: 624 SOP: MSV10 Rev: 15
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID:

Internal Standard: STD76580 Surrogate Standard: STD76383
CCV: STD76405 LCS: STD76744 MS/MSD: STD76744

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573776

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
11M12798	WG573773-01 50ng BFB STD 8260	NA	1	1	STD76496	06/22/16 14:52
11M12799	WG573773-02 50ug/L CCV STD 8260	NA	1	1	STD76405	06/22/16 15:16
11M12800	WG573773-02 50ug/L CCV STD 8260	NA	1	1	STD76405	06/22/16 15:49
11M12801	WG573773-02 50ug/L CCV STD 8260	NA	1	1	STD76405	06/22/16 16:43
11M12802	WGXXXXXX-01 100ug/L A9-CCV STD 826	NA	1	1	STDXXXXX	06/22/16 17:15
11M12803	WG573776-01 BLANK STD 8260	NA	1	1		06/22/16 17:47
11M12804	WG573776-02 20ug/L LCS STD 8260	NA	1	1	STD76744	06/22/16 18:19
11M12805	L16060974-09 A MS 2.5X 826-SPE	<2	1	2.5	STD76744	06/22/16 18:51
11M12806	L16060974-11 A MSD 2.5X 826-SPE	<2	1	2.5	STD76744	06/22/16 19:23
11M12807	L16060883-02 B D1 20X 826-SPE	<2	1	20		06/22/16 19:55
11M12808	L16061026-08 A TB 826-LOW	<2	1	1		06/22/16 20:27
11M12809	L16060974-01 A REF 2.5X 826-SPE	<2	1	2.5		06/22/16 20:59
11M12810	L16060974-03 A 826-SPE	<2	1	1		06/22/16 21:31
11M12811	L16060974-02 A 826-SPE	<2	1	1		06/22/16 22:03
11M12812	L16060974-04 A 826-SPE	<2	1	1		06/22/16 22:34
11M12813	L16061026-03 A 826-SPE	<2	1	1		06/22/16 23:06
11M12814	L16061026-07 A 826-SPE	<2	1	1		06/22/16 23:38
11M12815	L16061026-04 A 5X 826-LOW	<2	1	5		06/23/16 00:10
11M12816	L16061026-05 A 100X 826-LOW	<2	1	100		06/23/16 00:42
11M12817	L16061026-01 A 826-LOW	<2	1	1		06/23/16 01:14
11M12818	L16060974-05 A 826-SPE	<2	1	1		06/23/16 01:46
11M12819	L16060974-06 A 826-SPE	<2	1	1		06/23/16 02:18
11M12820	RINSE	NA	1	1		06/23/16 02:50
11M12821	WG573776-06 BLANK2 STD 624	NA	2	1		06/23/16 03:22
11M12822	L16061132-01 B 00 10X AF 624-SPE1	7	2	10		06/23/16 03:51
11M12823	L16061245-01 A 624-SPE	<2	2	1		06/23/16 04:23
11M12824	CCV	NA	1	1		06/23/16 04:55
11M12825	RINSE	NA	1	1		06/23/16 05:27
11M12826	RINSE	NA	1	1		06/23/16 05:59

Comments

Seq.	Rerun	Dil.	Reason	Analytes
3	X			

Approved: June 28, 2016

Sarah Vandenberg

Microbac

Page: 1

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS11 Dataset: 062216
Analyst1: JDS Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 24
Method: 624 SOP: MSV10 Rev: 15
Method: 5030B/5030C/5035A SOP: PAT01 Rev: 18
Maintenance Log ID:

Internal Standard: STD76580 Surrogate Standard: STD76383
CCV: STD76405 LCS: STD76744 MS/MSD: STD76744

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573776

Comments:

Comments

Seq.	Rerun	Dil.	Reason	Analytes
File ID: 11M12800				
	software error caused instrument data to be overwritten by snapshot data.			
8	X	1	Analyzed too dilute	
File ID: 11M12805				
	L16060974-09 MS			
9	X	1	Analyzed too dilute	
File ID: 11M12806				
	L16060974-11 MSD			
12	X	1	Analyzed too dilute	
File ID: 11M12809				
	L16060974-01 REF			
18	X	20	Over Calibration Range	trichloroethene
File ID: 11M12815				
	L16061026-04			
25				
File ID: 11M12822				
	L16061132-01 sample aborted by foam sensor. DNR			

Approved: June 28, 2016

Sarah Vandenberg

Page: 2

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS11 Dataset: 062316
Analyst1: JDS Analyst2: NA
Method: 8260B SOP: MSV01 Rev: 24
Method: 5030B/530C/5035A SOP: PAT01 Rev: 15

Maintenance Log ID: 53789

Internal Standard: STD76580 Surrogate Standard: STD76383
CCV: STD76781 LCS: STD76744 MS/MSD: STD76744

Column 1 ID: RTX502.2 Column 2 ID: NA
Workgroups: WG573953

Comments:

File ID	Sample Information	pH	Mat	Dil	Reference	Date/Time
11M12827	WG573951-01 50ng BFB STD 8260	NA	1	1	STD76496	06/23/16 15:00
11M12828	WG573951-02 50ug/L CCV STD 8260	NA	1	1	STD76781	06/23/16 15:25
11M12829	WG573951-02 50ug/L CCV STD 8260	NA	1	1	STD76781	06/23/16 15:59
11M12830	WGXXXXXX-01 100ug/L A9-CCV STD 826	NA	1	1	STDXXXXX	06/23/16 16:31
11M12831	WG573953-01 BLANK STD 8260	NA	1	1		06/23/16 17:03
11M12832	WG573953-02 20ug/L LCS STD 8260	NA	1	1	STD76744	06/23/16 17:35
11M12833	L16060974-09 B 00 MS 826-SPE	<2	1	1	STD76744	06/23/16 18:07
11M12834	L16060974-11 B 00 MSD 826-SPE	<2	1	1	STD76744	06/23/16 18:39
11M12835	L16061026-04 B D1 20X 826-LOW	<2	1	20		06/23/16 19:11
11M12836	L16060986-43 A TB 826-SPE	<2	1	1		06/23/16 19:43
11M12837	L16060974-01 B 00 REF 826-SPE	<2	1	1		06/23/16 20:15
11M12838	L16060986-39 A 826-SPE	<2	1	1		06/23/16 20:47
11M12839	L16060986-41 A 826-SPE	<2	1	1		06/23/16 21:19
11M12840	L16060986-44 A 826-SPE	<2	1	1		06/23/16 21:51
11M12841	L16060986-07 A 826-SPE	<2	1	1		06/23/16 22:22
11M12842	L16060986-09 A 826-SPE	<2	1	1		06/23/16 22:54
11M12843	L16060986-11 A 826-SPE	<2	1	1		06/23/16 23:26
11M12844	L16060986-17 A 826-SPE	<2	1	1		06/23/16 23:58
11M12845	L16060986-19 A 826-SPE	<2	1	1		06/24/16 00:30
11M12846	L16060986-25 A 826-SPE	<2	1	1		06/24/16 01:02
11M12847	L16060986-13 A 826-SPE	<2	1	1		06/24/16 01:34
11M12848	L16060986-27 A 826-SPE	<2	1	1		06/24/16 02:05
11M12849	L16060986-29 A 5X 826-SPE	<2	1	5		06/24/16 02:37
11M12850	CCV	NA	1	1		06/24/16 03:09
11M12851	RINSE	NA	1	1		06/24/16 03:41
11M12852	RINSE	NA	1	1		06/24/16 04:13

Approved: June 29, 2016
Sarah Vandenberg

Page: 1

Microbac Laboratories Inc.

Data Checklist

Date: 13-MAY-2016
Analyst: JDS
Analyst: NA
Method: 8260B
Instrument: HPMS11
Curve Workgroup: NA
Runlog ID: 75133
Analytical Workgroups: WG568769

System Performance Check	X
BFB	X
Initial Calibration	X
Average RF	X
Linear Reg or Higher Order Curve	X
Second Source standard % Difference	X
Continuing Calibration /Check Standards	X
Project/Client Specific Requirements	X
Special Standards	NA
Blanks	X
TCL's	X
Surrogates	X
LCS (Laboratory Control Sample)	X
Recoveries	X
Surrogates	X
MS/MSD/Duplicates	X
Samples	X
TCL Hits	X
Spectra of TCL Hits	X
Surrogates	X
Internal Standards Criteria	X
Library Searches	NA
Calculations & Correct Factors	X
Dilutions Run	X
Reruns	X
Manual Integrations	X
Case Narrative	X
Results Reporting/Data Qualifiers	X
KOBRA Workgroup Data	X
Check for Completeness	X
Primary Reviewer	JDS
Secondary Reviewer	FJB
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Check the reasonableness of the results	X

Primary Reviewer:
23-MAY-2016

Secondary Reviewer:
23-MAY-2016

CHECKLIST1 - Modified 03/05/2008
Generated: MAY-23-2016 15:39:32

Microbac Laboratories Inc.

Data Checklist

Date: 22-JUN-2016
Analyst: JDS
Analyst: NA
Method: 8260B
Instrument: HPMS11
Curve Workgroup: NA
Runlog ID: 75958
Analytical Workgroups: WG573776

System Performance Check	X
BFB	X
Initial Calibration	X
Average RF	X
Linear Reg or Higher Order Curve	X
Second Source standard % Difference	X
Continuing Calibration /Check Standards	X
Project/Client Specific Requirements	X
Special Standards	NA
Blanks	X
TCL's	X
Surrogates	X
LCS (Laboratory Control Sample)	X
Recoveries	X
Surrogates	X
MS/MSD/Duplicates	X
Samples	X
TCL Hits	X
Spectra of TCL Hits	WTD
Surrogates	X
Internal Standards Criteria	X
Library Searches	NA
Calculations & Correct Factors	X
Dilutions Run	X
Reruns	X
Manual Integrations	X
Case Narrative	X
Results Reporting/Data Qualifiers	X
KOBRA Workgroup Data	X
Check for Completeness	X
Primary Reviewer	JDS
Secondary Reviewer	SAV
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Check the reasonableness of the results	X

Primary Reviewer:
27-JUN-2016

Secondary Reviewer:
28-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-28-2016 12:52:44

Microbac Laboratories Inc.

Data Checklist

Date: 23-JUN-2016
Analyst: JDS
Analyst: NA
Method: 8260B
Instrument: HPMS11
Curve Workgroup: NA
Runlog ID: 76012
Analytical Workgroups: WG573953

System Performance Check	X
BFB	X
Initial Calibration	X
Average RF	X
Linear Reg or Higher Order Curve	X
Second Source standard % Difference	X
Continuing Calibration /Check Standards	X
Project/Client Specific Requirements	X
Special Standards	NA
Blanks	X
TCL's	X
Surrogates	X
LCS (Laboratory Control Sample)	X
Recoveries	X
Surrogates	X
MS/MSD/Duplicates	X
Samples	X
TCL Hits	X
Spectra of TCL Hits	WTD
Surrogates	X
Internal Standards Criteria	X
Library Searches	NA
Calculations & Correct Factors	X
Dilutions Run	X
Reruns	X
Manual Integrations	NA
Case Narrative	X
Results Reporting/Data Qualifiers	X
KOBRA Workgroup Data	X
Check for Completeness	X
Primary Reviewer	JDS
Secondary Reviewer	SAV
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Check the reasonableness of the results	X

Primary Reviewer:
29-JUN-2016

Secondary Reviewer:
29-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-29-2016 15:36:52

Microbac Laboratories Inc.

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 8260B

AAB#: WG573776

Login Number: L16061026

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18WW16-061616	01	06/16/16					06/23/2016	6.7	14		06/23/16	6.7	14	
102-061616	03	06/16/16					06/22/2016	6.6	14		06/22/16	6.6	14	
129-061616	04	06/16/16					06/23/2016	6.6	14		06/23/16	6.6	14	
120-061616	05	06/16/16					06/23/2016	6.6	14		06/23/16	6.6	14	
123-061616	07	06/16/16					06/22/2016	6.6	14		06/22/16	6.6	14	
TRIP BLANK	08	06/16/16					06/22/2016	6.9	14		06/22/16	6.9	14	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4827747
Report generated 06/29/2016 16:05

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 8260B

AAB#: WG573953

Login Number: L16061026

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
129-061616	04	06/16/16					06/23/2016	7.4	14		06/23/16	7.4	14	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4827747
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.
SURROGATE STANDARDS

Login Number: L16061026
Instrument Id: HPMS11
Workgroup (AAB#): WG573776

Method: 8260
CAL ID: HPMS11-13-MAY-16
Matrix: Water

Sample Number	Dilution	Tag	1	2	3	4
L16061026-01	1.00	01	104	108	91.3	100
L16061026-03	1.00	01	98.7	105	91.0	101
L16061026-04	5.00	DL01	100	106	94.7	102
L16061026-05	100	DL01	103	106	94.1	102
L16061026-07	1.00	01	99.4	106	93.5	100
L16061026-08	1.00	01	92.3	103	93.0	101
WG573776-01	1.00	01	93.4	101	92.6	99.9
WG573776-02	1.00	01	92.9	104	93.0	101
WG573776-06	1.00	01	105	107	95.5	101

Surrogates	Surrogate Limits		
1 - 1,2-Dichloroethane-d4	70	-	120
2 - Dibromofluoromethane	85	-	115
3 - 4-Bromofluorobenzene	75	-	120
4 - Toluene-d8	85	-	120

Underline = Result out of surrogate limits

DL = surrogate diluted out
ND = surrogate not detected

SURROGATES - Modified 03/06/2008
PDF File ID: 4827755
Report generated: 06/29/2016 16:05

Microbac Laboratories Inc.
SURROGATE STANDARDS

Login Number: L16061026
Instrument Id: HPMS11
Workgroup (AAB#): WG573953

Method: 8260
CAL ID: HPMS11-13-MAY-16
Matrix: Water

Sample Number	Dilution	Tag	1	2	3	4
L16061026-04	20.0	DL02	103	107	94.7	100
WG573953-01	1.00	01	100	104	93.8	100
WG573953-02	1.00	01	103	109	92.0	100

Surrogates	Surrogate Limits		
1 - 1,2-Dichloroethane-d4	70	-	120
2 - Dibromofluoromethane	85	-	115
3 - 4-Bromofluorobenzene	75	-	120
4 - Toluene-d8	85	-	120

Underline = Result out of surrogate limits

DL = surrogate diluted out
ND = surrogate not detected

SURROGATES - Modified 03/06/2008
PDF File ID: 4827755
Report generated: 06/29/2016 16:05

METHOD BLANK SUMMARY

Login Number: L16061026
Blank File ID: 11M12803
Prep Date: 06/22/16 17:47
Analyzed Date: 06/22/16 17:47
Analyst: JDS

Work Group: WG573776
Blank Sample ID: WG573776-01
Instrument ID: HPMS11
Method: 8260B

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573776-02	11M12804	06/22/16 18:19	01
TRIP BLANK	L16061026-08	11M12808	06/22/16 20:27	01
102-061616	L16061026-03	11M12813	06/22/16 23:06	01
123-061616	L16061026-07	11M12814	06/22/16 23:38	01
129-061616	L16061026-04	11M12815	06/23/16 00:10	DL01
120-061616	L16061026-05	11M12816	06/23/16 00:42	DL01
18WW16-061616	L16061026-01	11M12817	06/23/16 01:14	01

Report Name: BLANK_SUMMARY
PDF File ID: 4827748
Report generated 06/29/2016 16:05

METHOD BLANK SUMMARY

Login Number: L16061026
Blank File ID: 11M12831
Prep Date: 06/23/16 17:03
Analyzed Date: 06/23/16 17:03
Analyst: JDS

Work Group: WG573953
Blank Sample ID: WG573953-01
Instrument ID: HPMS11
Method: 8260B

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573953-02	11M12832	06/23/16 17:35	01
129-061616	L16061026-04	11M12835	06/23/16 19:11	DL02

Report Name: BLANK_SUMMARY
PDF File ID: 4827748
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.

METHOD BLANK REPORT

Login Number: L16061026 Prep Date: 06/22/16 17:47 Sample ID: WG573776-01
Instrument ID: HPMS11 Run Date: 06/22/16 17:47 Prep Method: 5030B/5030C/503
File ID: 11M12803 Analyst: JDS Method: 8260B
Workgroup (AAB#): WG573776 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS11-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Acetone	2.50	10.0	2.50	1	U
Benzene	0.125	1.00	0.125	1	U
Bromobenzene	0.125	1.00	0.125	1	U
Bromochloromethane	0.200	1.00	0.200	1	U
Bromodichloromethane	0.250	1.00	0.250	1	U
Bromoform	0.500	2.00	0.500	1	U
Bromomethane	0.500	2.00	0.500	1	U
2-Butanone	2.50	10.0	2.50	1	U
n-Butylbenzene	0.250	1.00	0.250	1	U
sec-Butylbenzene	0.250	1.00	0.250	1	U
tert-Butylbenzene	0.250	1.00	0.250	1	U
Carbon disulfide	0.500	2.00	0.500	1	U
Carbon tetrachloride	0.250	1.00	0.250	1	U
Chlorobenzene	0.125	1.00	0.125	1	U
Chlorodibromomethane	0.250	1.00	0.250	1	U
Chloroethane	0.500	2.00	0.500	1	U
Chloroform	0.125	1.00	0.125	1	U
Chloromethane	0.500	2.00	0.500	1	U
2-Chlorotoluene	0.125	1.00	0.125	1	U
4-Chlorotoluene	0.250	1.00	0.250	1	U
1,2-Dibromo-3-chloropropane	1.00	5.00	1.00	1	U
1,2-Dibromoethane	0.250	1.00	0.250	1	U
Dibromomethane	0.250	1.00	0.250	1	U
1,2-Dichlorobenzene	0.125	1.00	0.125	1	U
1,3-Dichlorobenzene	0.250	1.00	0.250	1	U
1,4-Dichlorobenzene	0.125	1.00	0.125	1	U
Dichlorodifluoromethane	0.250	1.00	0.250	1	U
1,1-Dichloroethane	0.125	1.00	0.125	1	U
1,2-Dichloroethane	0.250	1.00	0.250	1	U
1,1-Dichloroethene	0.500	2.00	0.500	1	U
cis-1,2-Dichloroethene	0.250	1.00	0.250	1	U
trans-1,2-Dichloroethene	0.250	1.00	0.250	1	U
1,2-Dichloropropane	0.200	1.00	0.200	1	U
1,3-Dichloropropane	0.200	1.00	0.200	1	U
2,2-Dichloropropane	0.250	1.00	0.250	1	U
cis-1,3-Dichloropropene	0.250	1.00	0.250	1	U
trans-1,3-Dichloropropene	0.500	2.00	0.500	1	U
1,1-Dichloropropene	0.250	1.00	0.250	1	U
Ethylbenzene	0.250	1.00	0.250	1	U
2-Hexanone	2.50	10.0	2.50	1	U
Hexachlorobutadiene	0.250	1.00	0.250	1	U
Isopropylbenzene	0.250	1.00	0.250	1	U

Report Name:BLANK
PDF ID: 4827749
29-JUN-2016 16:05

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16061026 Prep Date: 06/22/16 17:47 Sample ID: WG573776-01
Instrument ID: HPMS11 Run Date: 06/22/16 17:47 Prep Method: 5030B/5030C/503
File ID: 11M12803 Analyst: JDS Method: 8260B
Workgroup (AAB#): WG573776 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS11-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
p-Isopropyltoluene	0.250	1.00	0.250	1	U
4-Methyl-2-pentanone	2.50	10.0	2.50	1	U
Methylene chloride	0.250	1.00	0.250	1	U
Naphthalene	0.200	1.00	0.200	1	U
n-Propylbenzene	0.125	1.00	0.125	1	U
Styrene	0.125	1.00	0.125	1	U
1,1,1,2-Tetrachloroethane	0.250	1.00	0.250	1	U
1,1,2,2-Tetrachloroethane	0.200	1.00	0.200	1	U
Tetrachloroethene	0.250	1.00	0.250	1	U
Toluene	0.250	1.00	0.250	1	U
1,2,3-Trichlorobenzene	0.150	1.00	0.150	1	U
1,2,4-Trichlorobenzene	0.200	1.00	0.200	1	U
1,1,1-Trichloroethane	0.250	1.00	0.250	1	U
1,1,2-Trichloroethane	0.250	1.00	0.250	1	U
Trichloroethene	0.250	1.00	0.250	1	U
Trichlorofluoromethane	0.250	1.00	0.250	1	U
1,2,3-Trichloropropane	0.500	2.00	0.500	1	U
1,2,4-Trimethylbenzene	0.250	1.00	0.250	1	U
1,3,5-Trimethylbenzene	0.250	1.00	0.250	1	U
Vinyl chloride	0.250	1.00	0.250	1	U
o-Xylene	0.250	1.00	0.250	1	U
m-,p-Xylene	0.500	2.00	0.500	1	U

Surrogates	% Recovery	Surrogate Limits	Qualifier
Dibromofluoromethane	101	85 - 115	PASS
1,2-Dichloroethane-d4	93.4	70 - 120	PASS
Toluene-d8	99.9	85 - 120	PASS
4-Bromofluorobenzene	92.6	75 - 120	PASS

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4827749
29-JUN-2016 16:05

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16061026 Prep Date: 06/23/16 17:03 Sample ID: WG573953-01
Instrument ID: HPMS11 Run Date: 06/23/16 17:03 Prep Method: 5030B/5030C/503
File ID: 11M12831 Analyst: JDS Method: 8260B
Workgroup (AAB#): WG573953 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS11-13-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Trichloroethene	0.250	1.00	0.250	1	U

Surrogates	% Recovery	Surrogate Limits	Qualifier
Dibromofluoromethane	104	85 - 115	PASS
1,2-Dichloroethane-d4	100	70 - 120	PASS
Toluene-d8	100	85 - 120	PASS
4-Bromofluorobenzene	93.8	75 - 120	PASS

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name:BLANK
PDF ID: 4827749
29-JUN-2016 16:05

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16061026 | Run Date: 06/22/2016 | Sample ID: WG573776-02
Instrument ID: HPMS11 | Run Time: 18:19 | Prep Method: 5030B/5030C/503
File ID: 11M12804 | Analyst: JDS | Method: 8260B
Workgroup (AAB#): WG573776 | Matrix: Water | Units: ug/L
QC Key: DOD4 | Lot#: STD76744 | Cal ID: HPMS11-13-MAY-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Acetone	20.0	18.2	90.8	40 - 140	
Benzene	20.0	19.2	95.9	80 - 120	
Bromobenzene	20.0	19.8	98.8	75 - 125	
Bromochloromethane	20.0	22.4	112	65 - 130	
Bromodichloromethane	20.0	19.6	98.0	75 - 120	
Bromoform	20.0	19.8	98.9	70 - 130	
Bromomethane	20.0	23.4	117	30 - 145	
2-Butanone	20.0	19.3	96.6	30 - 150	
n-Butylbenzene	20.0	17.3	86.7	70 - 135	
sec-Butylbenzene	20.0	18.9	94.6	70 - 125	
tert-Butylbenzene	20.0	20.0	99.8	70 - 130	
Carbon disulfide	20.0	18.5	92.4	35 - 160	
Carbon tetrachloride	20.0	18.2	91.2	65 - 140	
Chlorobenzene	20.0	20.2	101	80 - 120	
Chlorodibromomethane	20.0	20.4	102	60 - 135	
Chloroethane	20.0	20.9	104	60 - 135	
Chloroform	20.0	18.6	93.1	65 - 135	
Chloromethane	20.0	20.3	102	40 - 125	
2-Chlorotoluene	20.0	19.5	97.6	75 - 125	
4-Chlorotoluene	20.0	18.9	94.5	75 - 130	
1,2-Dibromo-3-chloropropane	20.0	16.9	84.6	50 - 130	
1,2-Dibromoethane	20.0	20.8	104	80 - 120	
Dibromomethane	20.0	19.7	98.4	75 - 125	
1,2-Dichlorobenzene	20.0	20.3	101	70 - 120	
1,3-Dichlorobenzene	20.0	19.7	98.7	75 - 125	
1,4-Dichlorobenzene	20.0	19.9	99.3	75 - 125	
Dichlorodifluoromethane	20.0	21.8	109	30 - 155	
1,1-Dichloroethane	20.0	17.9	89.4	70 - 135	
1,2-Dichloroethane	20.0	19.0	95.2	70 - 130	
1,1-Dichloroethene	20.0	17.0	84.8	70 - 130	
cis-1,2-Dichloroethene	20.0	20.8	104	70 - 125	
trans-1,2-Dichloroethene	20.0	19.9	99.6	60 - 140	
1,2-Dichloropropane	20.0	19.3	96.3	75 - 125	
1,3-Dichloropropane	20.0	21.3	106	75 - 125	
2,2-Dichloropropane	20.0	17.8	89.2	70 - 135	
cis-1,3-Dichloropropene	20.0	21.8	109	70 - 130	
trans-1,3-Dichloropropene	20.0	19.5	97.5	55 - 140	
1,1-Dichloropropene	20.0	18.4	92.0	75 - 130	
Ethylbenzene	20.0	19.2	95.9	75 - 125	
2-Hexanone	20.0	16.4	82.0	55 - 130	
Hexachlorobutadiene	20.0	16.6	82.8	50 - 140	

LCS - Modified 03/06/2008
PDF File ID: 4827750
Report generated: 06/29/2016 16:05

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16061026 Run Date: 06/22/2016 Sample ID: WG573776-02
Instrument ID: HPMS11 Run Time: 18:19 Prep Method: 5030B/5030C/503
File ID: 11M12804 Analyst: JDS Method: 8260B
Workgroup (AAB#): WG573776 Matrix: Water Units: ug/L
QC Key: DOD4 Lot#: STD76744 Cal ID: HPMS11-13-MAY-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Isopropylbenzene	20.0	19.5	97.7	75 - 125	
p-Isopropyltoluene	20.0	18.8	94.1	75 - 130	
4-Methyl-2-pentanone	20.0	19.2	95.8	60 - 135	
Methylene chloride	20.0	20.4	102	55 - 140	
Naphthalene	20.0	18.1	90.5	55 - 140	
n-Propylbenzene	20.0	19.4	96.9	70 - 130	
Styrene	20.0	20.3	101	65 - 135	
1,1,1,2-Tetrachloroethane	20.0	19.7	98.5	80 - 130	
1,1,2,2-Tetrachloroethane	20.0	18.4	91.8	65 - 130	
Tetrachloroethene	20.0	19.3	96.7	45 - 150	
Toluene	20.0	19.5	97.7	75 - 120	
1,2,3-Trichlorobenzene	20.0	19.2	96.2	55 - 140	
1,2,4-Trichlorobenzene	20.0	19.5	97.5	65 - 135	
1,1,1-Trichloroethane	20.0	18.6	93.2	65 - 130	
1,1,2-Trichloroethane	20.0	20.9	104	75 - 125	
Trichloroethene	20.0	21.4	107	70 - 125	
Trichlorofluoromethane	20.0	18.6	92.9	60 - 145	
1,2,3-Trichloropropane	20.0	20.0	99.9	75 - 125	
1,2,4-Trimethylbenzene	20.0	19.3	96.5	75 - 130	
1,3,5-Trimethylbenzene	20.0	19.4	97.1	75 - 130	
Vinyl chloride	20.0	22.8	114	50 - 145	
o-Xylene	20.0	19.9	99.6	80 - 120	
m-,p-Xylene	40.0	38.9	97.3	75 - 130	

Surrogates	% Recovery	Surrogate Limits	Qualifier
Dibromofluoromethane	104	85 - 115	PASS
1,2-Dichloroethane-d4	92.9	70 - 120	PASS
Toluene-d8	101	85 - 120	PASS
4-Bromofluorobenzene	93.0	75 - 120	PASS

* EXCEEDS %REC LIMIT

LCS - Modified 03/06/2008
PDF File ID: 4827750
Report generated: 06/29/2016 16:05

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16061026 Run Date: 06/23/2016 Sample ID: WG573953-02
Instrument ID: HPMS11 Run Time: 17:35 Prep Method: 5030B/5030C/503
File ID: 11M12832 Analyst: JDS Method: 8260B
Workgroup (AAB#): WG573953 Matrix: Water Units: ug/L
QC Key: DOD4 Lot#: STD76744 Cal ID: HPMS11-13-MAY-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Trichloroethene	20.0	22.7	113	70 - 125	

Surrogates	% Recovery	Surrogate Limits	Qualifier
Dibromofluoromethane	109	85 - 115	PASS
1,2-Dichloroethane-d4	103	70 - 120	PASS
Toluene-d8	100	85 - 120	PASS
4-Bromofluorobenzene	92.0	75 - 120	PASS

* EXCEEDS %REC LIMIT

LCS - Modified 03/06/2008
PDF File ID: 4827750
Report generated: 06/29/2016 16:05

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

BFB

Login Number: L16061026
Instrument: HPMS11
Analyst: JDS
Workgroup: WG568769

Tune ID: WG568769-01
Run Date: 05/13/2016
Run Time: 14:15
File ID: 11M11836
Cal ID: HPMS11 - 13-MAY-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50.0	95.0	15.0	40.0	24.2	11638	PASS
75.0	95.0	30.0	60.0	50.2	24088	PASS
95.0	95.0	100	100	100	48016	PASS
96.0	95.0	5.00	9.00	6.26	3008	PASS
173	174	0	2.00	0	0	PASS
174	95.0	50.0	100	86.2	41402	PASS
175	174	5.00	9.00	8.26	3421	PASS
176	174	95.0	101	97.3	40288	PASS
177	176	5.00	9.00	6.92	2789	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG568769-02	STD	01	05/13/2016 14:40	
WG568769-03	STD	01	05/13/2016 15:12	
WG568769-04	STD	01	05/13/2016 15:43	
WG568769-05	STD	01	05/13/2016 16:15	
WG568769-06	STD	01	05/13/2016 16:47	
WG568769-07	STD	01	05/13/2016 17:19	
WG568769-08	STD-CCV	01	05/13/2016 17:51	
WG568769-09	STD	01	05/13/2016 18:22	
WG568769-10	STD	01	05/13/2016 18:54	
WG568769-11	STD	01	05/13/2016 19:26	
WG568769-12	SSCV	01	05/13/2016 20:30	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4827752
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

BFB

Login Number: L16061026
Instrument: HPMS11
Analyst: JDS
Workgroup: WG573773

Tune ID: WG573773-01
Run Date: 06/22/2016
Run Time: 14:52
File ID: 11M12798
Cal ID: HPMS11 - 13-MAY-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50.0	95.0	15.0	40.0	20.4	9816	PASS
75.0	95.0	30.0	60.0	48.3	23186	PASS
95.0	95.0	100	100	100	48053	PASS
96.0	95.0	5.00	9.00	7.38	3544	PASS
173	174	0	2.00	0	0	PASS
174	95.0	50.0	100	89.6	43074	PASS
175	174	5.00	9.00	7.57	3259	PASS
176	174	95.0	101	100	43058	PASS
177	176	5.00	9.00	6.83	2940	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573773-02	CCV	01	06/22/2016 16:43	
WG573776-01	BLANK	01	06/22/2016 17:47	
WG573776-02	LCS	01	06/22/2016 18:19	
L16061026-08	TRIP BLANK	01	06/22/2016 20:27	
L16061026-03	102-061616	01	06/22/2016 23:06	
L16061026-07	123-061616	01	06/22/2016 23:38	
L16061026-04	129-061616	DL01	06/23/2016 00:10	
L16061026-05	120-061616	DL01	06/23/2016 00:42	
L16061026-01	18WW16-061616	01	06/23/2016 01:14	
WG573776-06	BLANK2	01	06/23/2016 03:22	*

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4827752
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

BFB

Login Number: L16061026
Instrument: HPMS11
Analyst: JDS
Workgroup: WG573951

Tune ID: WG573951-01
Run Date: 06/23/2016
Run Time: 15:00
File ID: 11M12827
Cal ID: HPMS11 - 13-MAY-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50.0	95.0	15.0	40.0	21.6	6638	PASS
75.0	95.0	30.0	60.0	51.9	15962	PASS
95.0	95.0	100	100	100	30738	PASS
96.0	95.0	5.00	9.00	6.62	2035	PASS
173	174	0	2.00	0.406	109	PASS
174	95.0	50.0	100	87.4	26851	PASS
175	174	5.00	9.00	7.85	2109	PASS
176	174	95.0	101	95.6	25668	PASS
177	176	5.00	9.00	7.02	1803	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573951-02	CCV	01	06/23/2016 15:59	
WG573953-01	BLANK	01	06/23/2016 17:03	
WG573953-02	LCS	01	06/23/2016 17:35	
L16061026-04	129-061616	DL02	06/23/2016 19:11	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4827752
Report generated 06/29/2016 16:05

Calibration Table Report
Method: 8260WT.M
Title: 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Calibration: Sat May 14 18:45:57 2016
Curve: WG568769
Calibration Files

	Compound	0.3 11M11837.D	0.4 11M11838.D	1 11M11839.D	2 11M11840.D	5 11M11841.D	20 11M11842.D	50 11M11843.D	100 11M11844.D	200 11M11845.D	300 11M11846.D	Avg	%RSD	Linear	Quadratic
I	Fluorobenzene	ISTD													
T	Dichlorodifluoromethane			0.345	0.368	0.326	0.431	0.400	0.424	0.417	0.424	0.392	10.308		
P	Chloromethane			0.348	0.363	0.289	0.310	0.276	0.296	0.305	0.324	0.314	9.474		
C	Vinyl Chloride		0.275	0.241	0.249	0.244	0.268	0.250	0.271	0.271	0.276	0.261	5.458		
T	1,3-Butadiene					0.242	0.284	0.269	0.264	0.256	0.240	0.259	6.530		
T	Bromomethane			0.167	0.174	0.164	0.180	0.165	0.182	0.194	0.206	0.179	8.292		
T	Chloroethane			0.164	0.165	0.151	0.174	0.161	0.171	0.173	0.181	0.167	5.524		
T	Trichlorofluoromethane		0.514	0.523	0.537	0.496	0.552	0.498	0.524	0.514	0.514	0.519	3.376		
T	Diethyl ether			0.189	0.191	0.195	0.196	0.188	0.193		0.192	0.192	1.552		
T	Isoprene					0.299	0.331	0.317	0.323	0.325	0.325	0.320	3.522		
T	Acrolein			0.013	0.022	0.024	0.024	0.024	0.025		0.027	0.023	19.436	0.999	
T	1,1,2-Trichloro-1,2,2-Trifluor			0.261	0.245	0.234	0.268	0.245	0.252	0.254	0.257	0.252	4.225		
T	Acetone					0.063	0.067	0.058	0.057	0.057	0.054	0.059	7.931		
C	1,1-Dichloroethene		0.459	0.452	0.460	0.429	0.492	0.441	0.464	0.461	0.461	0.458	3.810		
T	Tert-Butyl Alcohol			0.014	0.015	0.017	0.018	0.017	0.017		0.017	0.016	9.673		
T	Dimethyl Sulfide					0.169	0.185	0.186	0.185	0.188	0.184	0.183	3.765		
T	Iodomethane				0.043	0.072	0.164	0.206	0.217	0.219	0.208	0.162	45.603		0.999
T	Methyl acetate				0.182	0.317	0.215	0.178	0.176	0.169	0.168	0.201	26.745		1.000
T	Methylene Chloride			0.239	0.256	0.230	0.243	0.224	0.232	0.234	0.238	0.237	4.076		
T	Carbon Disulfide			0.737	0.762	0.714	0.798	0.783	0.783	0.782	0.770	0.766	3.624		
T	Acrylonitrile			0.072	0.069	0.075	0.085	0.081	0.085		0.090	0.080	9.576		
T	Methyl Tert Butyl Ether			0.568	0.546	0.564	0.650	0.585	0.606	0.612	0.595	0.591	5.543		
T	trans-1,2-Dichloroethene		0.288	0.230	0.264	0.229	0.263	0.242	0.251	0.254	0.260	0.253	7.263		
T	n-Hexane				0.424	0.377	0.450	0.431	0.426	0.424	0.422	0.422	5.186		
T	Diisopropyl ether			1.034	1.075	1.083	1.055	1.046	1.062		1.015	1.053	2.218		
T	Vinyl Acetate					0.235	0.391	0.393	0.395	0.430	0.431	0.379	19.260	0.999	
P	1,1-Dichloroethane		0.463	0.517	0.527	0.505	0.535	0.495	0.513	0.511	0.508	0.508	4.023		
T	Ethyl-Tert-Butyl ether			0.815	0.854	0.881	0.876	0.855	0.873		0.844	0.857	2.667		
T	2-Butanone					0.079	0.097	0.087	0.092	0.092	0.091	0.090	6.720		
T	Propionitrile			0.023	0.025	0.026	0.029	0.027	0.027		0.027	0.026	7.381		
T	2,2-Dichloropropane		0.325	0.372	0.435	0.395	0.406	0.386	0.422	0.418	0.419	0.398	8.479		
T	cis-1,2-Dichloroethene		0.276	0.266	0.289	0.269	0.294	0.271	0.280	0.284	0.285	0.279	3.456		
C	Chloroform	0.587	0.483	0.491	0.514	0.480	0.509	0.470	0.486	0.489	0.482	0.499	6.767		
T	1-Bromopropane			0.014	0.041	0.045	0.047	0.048	0.049	0.050	0.050	0.043	28.190	1.000	
T	Bromochloromethane		0.115	0.140	0.169	0.167	0.188	0.172	0.179	0.180	0.181	0.166	14.065		
T	Tetrahydrofuran			0.074	0.061	0.058	0.067	0.061	0.062		0.059	0.063	8.769		
S	Dibromofluoromethane				0.251	0.266	0.284	0.268	0.284	0.285	0.278	0.274	4.605		
T	1,1,1-Trichloroethane		0.475	0.478	0.508	0.471	0.515	0.477	0.502	0.499	0.495	0.491	3.293		
T	Cyclohexane		0.477	0.485	0.547	0.509	0.574	0.551	0.560	0.558	0.556	0.535	6.606		
T	1,1-Dichloropropene		0.342	0.350	0.346	0.332	0.370	0.339	0.358	0.359	0.362	0.351	3.562		
T	Carbon Tetrachloride		0.409	0.457	0.522	0.465	0.518	0.473	0.496	0.492	0.486	0.480	7.177		
T	Tert-Amyl-Methyl ether			0.565	0.596	0.603	0.611	0.588	0.605		0.6	0.59537	2.5818		
S	1,2-Dichloroethane-d4			0.315	0.304	0.32	0.326	0.303	0.319	0.31	0.298	0.31197	3.15125		
T	1,2-Dichloroethane		0.388	0.391	0.419	0.395	0.441	0.402	0.413	0.409	0.391	0.40536	4.2274		
T	Benzene		1.058	0.952	1.016	0.923	0.996	0.908	0.943	0.931	0.914	0.96002	5.38884		
T	Trichloroethene		0.308	0.317	0.329	0.306	0.322	0.294	0.313	0.309	0.311	0.31204	3.21019		
T	Methylcyclohexane		0.332	0.37	0.389	0.35	0.411	0.401	0.4	0.405	0.407	0.3851	7.28708		
C	1,2-Dichloropropane		0.259	0.262	0.273	0.254	0.276	0.257	0.268	0.27	0.273	0.26575	2.95636		
T	1,4-Dioxane				0.001	0.001	0.002	0.001	0.002		0.002	0.0014	19.4617	0.999	
T	Bromodichloromethane		0.346	0.346	0.373	0.362	0.4	0.369	0.387	0.389	0.385	0.37307	5.09783		
T	Dibromomethane		0.118	0.136	0.135	0.137	0.149	0.137	0.146	0.145	0.144	0.13848	6.69192		
T	2-Chloroethyl Vinyl Ether					0.112	0.138	0.128	0.136	0.138	0.135	0.13098	7.7176		
T	4-Methyl-2-Pentanone					0.06	0.08	0.073	0.076	0.077	0.077	0.07387	9.67651		
T	cis-1,3-Dichloropropene		0.326	0.338	0.358	0.357	0.406	0.379	0.395	0.401	0.399	0.37315	7.91395		
T	Dimethyl Disulfide					0.185	0.229	0.227	0.233	0.238	0.235	0.22454	8.74309		

I	Chlorobenzene-d5	ISTD												
S	Toluene-d8				1.101	1.093	1.115	1.056	1.103	1.092	1.044	1.08625	2.42896	
C	Toluene		1.309	1.253	1.287	1.269	1.347	1.249	1.287	1.241	1.149	1.2656	4.33323	
T	Ethyl Methacrylate			0.234	0.233	0.25	0.299	0.285	0.289	0.291	0.281	0.27024	9.93745	
T	trans-1,3-Dichloropropene			0.335	0.388	0.387	0.448	0.418	0.432	0.43	0.412	0.40623	8.82781	
T	1,1,2-Trichloroethane		0.222	0.201	0.243	0.21	0.246	0.221	0.228	0.227	0.221	0.22442	6.30977	
T	2-Hexanone					0.137	0.179	0.165	0.172	0.172	0.165	0.16482	8.91224	
T	1,3-Dichloropropane		0.337	0.338	0.351	0.355	0.396	0.361	0.374	0.373	0.358	0.36039	5.16222	
T	Tetrachloroethene		0.274	0.287	0.317	0.286	0.306	0.286	0.3	0.299	0.294	0.29431	4.41886	
T	Dibromochloromethane		0.283	0.297	0.336	0.33	0.378	0.354	0.372	0.372	0.359	0.3423	9.88235	
T	1,2-Dibromoethane		0.232	0.182	0.228	0.214	0.243	0.224	0.233	0.235	0.232	0.22485	7.92714	
T	1-Chlorohexane		0.37	0.399	0.414	0.391	0.442	0.429	0.44	0.438	0.432	0.41726	6.09034	
P	Chlorobenzene		0.927	0.917	0.998	0.932	0.976	0.909	0.93	0.902	0.853	0.92715	4.48215	
T	1,1,1,2-Tetrachloroethane		0.341	0.361	0.405	0.366	0.388	0.371	0.384	0.385	0.376	0.37515	4.87731	
C	Ethylbenzene		0.453	0.457	0.512	0.463	0.495	0.466	0.482	0.479	0.472	0.47542	3.94078	
T	m-,p-Xylene		0.619	0.583	0.605	0.564	0.604	0.568	0.579	0.559	0.522	0.57798	5.08853	
T	o-Xylene		0.539	0.549	0.574	0.539	0.6	0.561	0.574	0.57	0.557	0.56249	3.43829	
T	Styrene		0.865	0.848	0.897	0.918	1.018	0.955	0.985	0.959	0.917	0.92933	5.96246	
P	Bromoform			0.172	0.189	0.185	0.226	0.21	0.225	0.226	0.228	0.20747	10.9157	
T	Isopropylbenzene		1.515	1.456	1.567	1.46	1.592	1.493	1.534	1.449	1.323	1.48754	5.34769	
I	1,4-Dichlorobenzene-d4	ISTD												
P	1,1,2,2-Tetrachloroethane		0.288	0.357	0.369	0.371	0.441	0.396	0.412	0.428	0.442	0.38931	12.6902	
S	p-Bromofluorobenzene				0.772	0.784	0.794	0.716	0.767	0.789	0.773	0.77065	3.38935	
T	1,2,3-Trichloropropane			0.087	0.129	0.138	0.151	0.137	0.142	0.148	0.148	0.13515	15.3522	0.999
T	trans-1,4-Dichloro-2-Butene			0.105	0.117	0.124	0.177	0.178	0.184	0.192	0.193	0.15883	23.0996	0.999
T	n-Propylbenzene		2.87	2.8	3.077	2.949	3.134	2.886	2.979	2.836	2.526	2.89527	6.10377	
T	Bromobenzene	0.663	0.778	0.81	0.798	0.765	0.784	0.718	0.752	0.769	0.76	0.75973	5.57443	
T	1,3,5-Trimethylbenzene		2.338	2.056	2.263	2.172	2.357	2.166	2.252	2.223	2.059	2.20939	4.86261	
T	2-Chlorotoluene		2.158	2.096	2.122	2.033	2.128	1.936	2.002	1.954	1.82	2.0279	5.4704	
T	4-Chlorotoluene		1.782	1.755	1.937	1.808	1.916	1.763	1.832	1.838	1.717	1.8165	4.03452	
T	a-Methylstyrene			1	1.066	1.101	1.241	1.231	1.235	1.242	1.208	1.16555	8.18507	
T	tert-Butylbenzene		0.378	0.44	0.48	0.489	0.498	0.458	0.483	0.49	0.497	0.46809	8.31778	
T	1,2,4-Trimethylbenzene		2.159	2.195	2.274	2.295	2.472	2.234	2.309	2.244	2.067	2.24981	4.97706	
T	sec-Butylbenzene			2.593	2.734	2.647	2.87	2.634	2.723	2.617	2.404	2.65291	5.05975	
T	p-Isopropyltoluene			2.247	2.592	2.45	2.641	2.43	2.506	2.404	2.226	2.43674	6.05243	
T	1,3-Dichlorobenzene		1.487	1.488	1.538	1.46	1.53	1.398	1.451	1.446	1.406	1.46732	3.33475	
T	1,4-Dichlorobenzene	1.482	1.601	1.511	1.569	1.484	1.538	1.404	1.456	1.434	1.395	1.4873	4.59741	
T	n-Butylbenzene		2.048	2.099	2.295	2.049	2.32	2.134	2.224	2.136	2.021	2.14721	5.09289	
T	1,2-Dichlorobenzene	1.33	1.26	1.333	1.432	1.33	1.406	1.296	1.343	1.326	1.31	1.33665	3.71841	
T	1,2-Dibromo-3-Chloropropane				0.057	0.079	0.096	0.081	0.086	0.087	0.091	0.08248	15.3444	0.999
T	1,2,4-Trichlorobenzene		0.927	0.903	0.957	0.956	1.042	0.972	0.998	1.002	1.047	0.97808	4.96993	
T	Hexachlorobutadiene		0.393	0.381	0.451	0.418	0.447	0.415	0.441	0.442	0.49	0.43089	7.64497	
T	Naphthalene		1.885	1.633	1.765	1.712	2.026	1.824	1.878	1.821	1.794	1.81543	6.15168	
T	1,2,3-Trichlorobenzene	0.854	0.957	0.865	0.918	0.834	0.946	0.86	0.889	0.912	0.943	0.89764	4.8559	

Tue May 17 13:57:02 2016

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16061026 Run Date: 05/13/2016 Sample ID: WG568769-12
Instrument ID: HPMS11 Run Time: 20:30 Method: 8260B
File ID: 11M11848 Analyst: JDS QC Key: DOD4
ICal Workgroup: WG568769 Cal ID: HPMS11 - 13-MAY-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
Chloroform	CCC	50.0	47.6	ug/L	0.475	4.80	20	
1,1-Dichloroethene	CCC	50.0	47.5	ug/L	0.435	4.90	20	
1,2-Dichloropropane	CCC	50.0	52.8	ug/L	0.281	5.60	20	
Ethylbenzene	CCC	50.0	52.4	ug/L	0.498	4.80	20	
Toluene	CCC	50.0	53.8	ug/L	1.36	7.50	20	
Vinyl Chloride	CCC	50.0	56.1	ug/L	0.292	12.2	20	
Bromoform	SPCC	50.0	49.8	ug/L	0.207	0.300	20	
Chlorobenzene	SPCC	50.0	53.3	ug/L	0.988	6.60	20	
Chloromethane	SPCC	50.0	55.1	ug/L	0.346	10.2	20	
1,1-Dichloroethane	SPCC	50.0	48.5	ug/L	0.493	3.00	20	
1,1,2,2-Tetrachloroethane	SPCC	50.0	54.9	ug/L	0.428	9.90	20	
Acetone		50.0	49.7	ug/L	0.0590	0.700	20	
Benzene		50.0	51.2	ug/L	0.983	2.40	20	
Bromobenzene		50.0	51.6	ug/L	0.785	3.30	20	
Bromochloromethane		50.0	53.9	ug/L	0.179	7.80	20	
Bromodichloromethane		50.0	49.5	ug/L	0.370	0.900	20	
Bromomethane		50.0	50.5	ug/L	0.181	1.00	20	
2-Butanone		50.0	50.3	ug/L	0.0904	0.700	20	
n-Butylbenzene		50.0	53.9	ug/L	2.31	7.80	20	
sec-Butylbenzene		50.0	55.9	ug/L	2.97	11.8	20	
tert-Butylbenzene		50.0	56.5	ug/L	0.529	13.1	20	
Carbon Disulfide		50.0	44.7	ug/L	0.685	10.6	20	
Carbon Tetrachloride		50.0	47.3	ug/L	0.454	5.30	20	
Dibromochloromethane		50.0	52.5	ug/L	0.360	5.10	20	
Chloroethane		50.0	58.9	ug/L	0.197	17.7	20	
2-Chlorotoluene		50.0	52.8	ug/L	2.14	5.50	20	
4-Chlorotoluene		50.0	55.4	ug/L	2.01	10.8	20	
1,2-Dibromo-3-Chloropropane		50.0	48.1	ug/L	0.0844	3.80	20	
1,2-Dibromoethane		50.0	52.6	ug/L	0.237	5.30	20	
Dibromomethane		50.0	48.0	ug/L	0.133	3.90	20	
1,2-Dichlorobenzene		50.0	53.6	ug/L	1.43	7.20	20	
1,3-Dichlorobenzene		50.0	52.7	ug/L	1.55	5.40	20	
1,4-Dichlorobenzene		50.0	53.0	ug/L	1.58	6.10	20	
Dichlorodifluoromethane		50.0	56.6	ug/L	0.443	13.2	20	
1,2-Dichloroethane		50.0	47.9	ug/L	0.388	4.20	20	
cis-1,2-Dichloroethene		50.0	53.2	ug/L	0.297	6.30	20	
trans-1,2-Dichloroethene		50.0	52.1	ug/L	0.264	4.30	20	
1,3-Dichloropropane		50.0	56.3	ug/L	0.406	12.6	20	
2,2-Dichloropropane		50.0	45.6	ug/L	0.362	8.90	20	
cis-1,3-Dichloropropene		50.0	57.6	ug/L	0.430	15.2	20	
trans-1,3-Dichloropropene		50.0	52.1	ug/L	0.423	4.20	20	
1,1-Dichloropropene		50.0	50.1	ug/L	0.352	0.200	20	

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4827751
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16061026 Run Date: 05/13/2016 Sample ID: WG568769-12
Instrument ID: HPMS11 Run Time: 20:30 Method: 8260B
File ID: 11M11848 Analyst: JDS QC Key: DOD4
ICal Workgroup: WG568769 Cal ID: HPMS11 - 13-MAY-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
2-Hexanone		50.0	52.2	ug/L	0.172	4.40	20	
Hexachlorobutadiene		50.0	55.9	ug/L	0.482	11.8	20	
Isopropylbenzene		50.0	54.2	ug/L	1.61	8.50	20	
p-Isopropyltoluene		50.0	54.7	ug/L	2.67	9.40	20	
4-Methyl-2-Pentanone		50.0	52.6	ug/L	0.0777	5.10	20	
Methylene Chloride		50.0	52.0	ug/L	0.246	4.00	20	
Naphthalene		50.0	51.1	ug/L	1.86	2.30	20	
n-Propylbenzene		50.0	56.6	ug/L	3.28	13.2	20	
Styrene		50.0	55.0	ug/L	1.02	9.90	20	
1,1,1,2-Tetrachloroethane		50.0	51.0	ug/L	0.383	2.10	20	
Tetrachloroethene		50.0	52.2	ug/L	0.307	4.40	20	
1,2,3-Trichlorobenzene		50.0	54.1	ug/L	0.971	8.10	20	
1,2,4-Trichlorobenzene		50.0	55.4	ug/L	1.08	10.9	20	
1,1,1-Trichloroethane		50.0	48.1	ug/L	0.472	3.90	20	
1,1,2-Trichloroethane		50.0	52.8	ug/L	0.237	5.50	20	
Trichloroethene		50.0	52.3	ug/L	0.327	4.60	20	
Trichlorofluoromethane		50.0	47.5	ug/L	0.494	4.90	20	
1,2,3-Trichloropropane		50.0	50.1	ug/L	0.147	0.300	20	
1,2,4-Trimethylbenzene		50.0	53.8	ug/L	2.42	7.70	20	
1,3,5-Trimethylbenzene		50.0	55.1	ug/L	2.44	10.3	20	
o-Xylene		50.0	54.7	ug/L	0.615	9.40	20	
m-,p-Xylene		100	107	ug/L	0.620	7.20	20	

* Exceeds %D Limit

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4827751
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.
CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026 Run Date: 06/22/2016 Sample ID: WG573773-02
Instrument ID: HPMS11 Run Time: 16:43 Method: 8260B
File ID: 11M12801 Analyst: JDS QC Key: DOD4
Workgroup (AAB#): WG573776 Cal ID: HPMS11 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Chloroform	CCC	50.0	48.8	ug/L	0.488	2.32	20	
1,1-Dichloroethene	CCC	50.0	49.6	ug/L	0.454	0.854	20	
1,2-Dichloropropane	CCC	50.0	49.9	ug/L	0.265	0.160	20	
Ethylbenzene	CCC	50.0	52.1	ug/L	0.495	4.11	20	
Toluene	CCC	50.0	52.7	ug/L	1.33	5.41	20	
Vinyl Chloride	CCC	50.0	56.5	ug/L	0.294	12.9	20	
Bromoform	SPCC	50.0	53.8	ug/L	0.223	7.54	20	
Chlorobenzene	SPCC	50.0	52.1	ug/L	0.966	4.14	20	
Chloromethane	SPCC	50.0	51.6	ug/L	0.324	3.17	20	
1,1-Dichloroethane	SPCC	50.0	49.9	ug/L	0.507	0.198	20	
1,1,2,2-Tetrachloroethane	SPCC	50.0	50.7	ug/L	0.395	1.35	20	
Xylenes		150	156	ug/L	0.592	3.81	20	
Acetone		50.0	40.8	ug/L	0.0484	18.5	20	
Benzene		50.0	51.2	ug/L	0.984	2.45	20	
Bromobenzene		50.0	52.3	ug/L	0.795	4.58	20	
Bromochloromethane		50.0	59.1	ug/L	0.196	18.2	20	
Bromodichloromethane		50.0	50.8	ug/L	0.379	1.57	20	
Bromomethane		50.0	60.5	ug/L	0.217	21.0	20	*
2-Butanone		50.0	41.0	ug/L	0.0735	18.1	20	
n-Butylbenzene		50.0	48.5	ug/L	2.08	2.97	20	
sec-Butylbenzene		50.0	50.7	ug/L	2.69	1.38	20	
tert-Butylbenzene		50.0	52.9	ug/L	0.495	5.77	20	
Carbon Disulfide		50.0	53.9	ug/L	0.826	7.79	20	
Carbon Tetrachloride		50.0	50.7	ug/L	0.486	1.33	20	
Dibromochloromethane		50.0	54.3	ug/L	0.372	8.63	20	
Chloroethane		50.0	54.4	ug/L	0.182	8.85	20	
2-Chlorotoluene		50.0	49.7	ug/L	2.02	0.598	20	
4-Chlorotoluene		50.0	49.3	ug/L	1.79	1.39	20	
1,2-Dibromo-3-Chloropropane		50.0	44.1	ug/L	0.0773	11.8	20	
1,2-Dibromoethane		50.0	52.5	ug/L	0.236	5.06	20	
Dibromomethane		50.0	51.4	ug/L	0.142	2.79	20	
1,2-Dichlorobenzene		50.0	50.8	ug/L	1.36	1.54	20	
1,3-Dichlorobenzene		50.0	51.3	ug/L	1.51	2.57	20	
1,4-Dichlorobenzene		50.0	50.4	ug/L	1.50	0.809	20	
Dichlorodifluoromethane		50.0	56.3	ug/L	0.441	12.7	20	
1,2-Dichloroethane		50.0	47.1	ug/L	0.382	5.81	20	
cis-1,2-Dichloroethene		50.0	53.4	ug/L	0.298	6.81	20	
trans-1,2-Dichloroethene		50.0	54.1	ug/L	0.274	8.15	20	
1,3-Dichloropropane		50.0	51.2	ug/L	0.369	2.33	20	
2,2-Dichloropropane		50.0	53.9	ug/L	0.429	7.88	20	
cis-1,3-Dichloropropene		50.0	53.4	ug/L	0.399	6.84	20	
trans-1,3-Dichloropropene		50.0	51.7	ug/L	0.420	3.38	20	

CCV - Modified 03/05/2008
PDF File ID: 4827753
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026 Run Date: 06/22/2016 Sample ID: WG573773-02
Instrument ID: HPMS11 Run Time: 16:43 Method: 8260B
File ID: 11M12801 Analyst: JDS QC Key: DOD4
Workgroup (AAB#): WG573776 Cal ID: HPMS11 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
1,1-Dichloropropene		50.0	51.8	ug/L	0.364	3.68	20	
2-Hexanone		50.0	38.7	ug/L	0.127	22.7	20	*
Hexachlorobutadiene		50.0	49.5	ug/L	0.427	0.931	20	
Isopropylbenzene		50.0	51.2	ug/L	1.52	2.34	20	
p-Isopropyltoluene		50.0	51.0	ug/L	2.48	1.91	20	
4-Methyl-2-Pentanone		50.0	44.7	ug/L	0.0660	10.6	20	
Methylene Chloride		50.0	51.4	ug/L	0.244	2.89	20	
Naphthalene		50.0	50.1	ug/L	1.82	0.171	20	
n-Propylbenzene		50.0	51.1	ug/L	2.96	2.26	20	
Styrene		50.0	54.0	ug/L	1.00	7.94	20	
1,1,1,2-Tetrachloroethane		50.0	51.9	ug/L	0.390	3.90	20	
Tetrachloroethene		50.0	54.5	ug/L	0.321	9.08	20	
1,2,3-Trichlorobenzene		50.0	48.9	ug/L	0.878	2.20	20	
1,2,4-Trichlorobenzene		50.0	50.6	ug/L	0.990	1.23	20	
1,1,1-Trichloroethane		50.0	50.1	ug/L	0.493	0.290	20	
1,1,2-Trichloroethane		50.0	52.0	ug/L	0.234	4.06	20	
Trichloroethene		50.0	53.9	ug/L	0.337	7.86	20	
Trichlorofluoromethane		50.0	51.3	ug/L	0.533	2.70	20	
1,2,3-Trichloropropane		50.0	49.1	ug/L	0.143	1.82	20	
1,2,4-Trimethylbenzene		50.0	51.1	ug/L	2.30	2.11	20	
1,3,5-Trimethylbenzene		50.0	50.9	ug/L	2.25	1.81	20	
o-Xylene		50.0	51.8	ug/L	0.582	3.51	20	
m-,p-Xylene		100	104	ug/L	0.601	3.95	20	

* Exceeds %D Criteria

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

CCV - Modified 03/05/2008
PDF File ID: 4827753
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/23/2016
Sample ID: WG573951-02
Instrument ID: HPMS11
Run Time: 15:59
Method: 8260B
File ID: 11M12829
Analyst: JDS
QC Key: DOD4
Workgroup (AAB#): WG573953
Cal ID: HPMS11 - 13-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Chloroform	CCC	50.0	50.3	ug/L	0.502	0.640	20	
1,1-Dichloroethene	CCC	50.0	50.0	ug/L	0.458	0.000800	20	
1,2-Dichloropropane	CCC	50.0	49.0	ug/L	0.261	1.97	20	
Ethylbenzene	CCC	50.0	49.1	ug/L	0.467	1.79	20	
Toluene	CCC	50.0	49.3	ug/L	1.25	1.50	20	
Vinyl Chloride	CCC	50.0	57.5	ug/L	0.300	15.0	20	
Bromoform	SPCC	50.0	54.2	ug/L	0.225	8.31	20	
Chlorobenzene	SPCC	50.0	50.1	ug/L	0.929	0.231	20	
Chloromethane	SPCC	50.0	53.2	ug/L	0.334	6.45	20	
1,1-Dichloroethane	SPCC	50.0	49.6	ug/L	0.504	0.761	20	
1,1,2,2-Tetrachloroethane	SPCC	50.0	46.5	ug/L	0.362	6.94	20	
Trichloroethene		50.0	52.3	ug/L	0.326	4.56	20	

* Exceeds %D Criteria

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

CCV - Modified 03/05/2008
PDF File ID: 4827753
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16061026
Instrument ID: HPMS11
Workgroup (AAB#): WG573776

ICAL CCV Number: WG568769-08
CAL ID: HPMS11-13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568769-08	NA	NA	263885	438582	521583
Upper Limit	NA	NA	527770	877164	1043166
Lower Limit	NA	NA	131943	219291	260792
L16061026-01	1.00	01	262492	442694	496621
L16061026-03	1.00	01	286068	484683	555060
L16061026-04	5.00	DL01	276226	471668	534594
L16061026-05	100	DL01	269312	458713	517460
L16061026-07	1.00	01	285222	492737	557279
L16061026-08	1.00	01	313293	533935	611363
WG573776-01	1.00	01	320378	558566	634960
WG573776-02	1.00	01	335580	569515	650667

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_ICAL - Modified 03/06/2008
PDF File ID: 4827754
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16061026
Instrument ID: HPMS11
Workgroup (AAB#): WG573953

ICAL CCV Number: WG568769-08
CAL ID: HPMS11-13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568769-08	NA	NA	263885	438582	521583
Upper Limit	NA	NA	527770	877164	1043166
Lower Limit	NA	NA	131943	219291	260792
L16061026-04	20.0	DL02	268568	457128	507627
WG573953-01	1.00	01	257668	439921	499965
WG573953-02	1.00	01	275615	445760	489086

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_ICAL - Modified 03/06/2008
PDF File ID: 4827754
Report generated 06/29/2016 16:05

Microbac Laboratories Inc.
INTERNAL STANDARD RETENTION TIME SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16061026
Instrument ID: HPMS11
Workgroup (AAB#): WG573776

ICAL CCV Number: WG568769-08
CAL ID: HPMS11-13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568769-08	NA	NA	17.06	14.25	10.61
Upper Limit	NA	NA	17.56	14.75	11.11
Lower Limit	NA	NA	16.56	13.75	10.11
L16061026-01	1.00	01	17.06	14.25	10.61
L16061026-03	1.00	01	17.06	14.25	10.61
L16061026-04	5.00	DL01	17.06	14.25	10.61
L16061026-05	100	DL01	17.06	14.25	10.61
L16061026-07	1.00	01	17.06	14.25	10.61
L16061026-08	1.00	01	17.06	14.25	10.61
WG573776-01	1.00	01	17.06	14.25	10.61
WG573776-02	1.00	01	17.06	14.25	10.61

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_RT_ICAL - Modified 03/06/2008
PDF File ID: 4827756
Report generated: 06/29/2016 16:05

Microbac Laboratories Inc.

INTERNAL STANDARD RETENTION TIME SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16061026
Instrument ID: HPMS11
Workgroup (AAB#): WG573953

ICAL CCV Number: WG568769-08
CAL ID: HPMS11-13-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1	IS-2	IS-3
WG568769-08	NA	NA	17.06	14.25	10.61
Upper Limit	NA	NA	17.56	14.75	11.11
Lower Limit	NA	NA	16.56	13.75	10.11
L16061026-04	20.0	DL02	17.06	14.25	10.61
WG573953-01	1.00	01	17.06	14.25	10.61
WG573953-02	1.00	01	17.06	14.25	10.61

IS-1 - 1,4-Dichlorobenzene-d4
IS-2 - Chlorobenzene-d5
IS-3 - Fluorobenzene

Underline = Response outside limits

INTERNAL_STD_RT_ICAL - Modified 03/06/2008
PDF File ID: 4827756
Report generated: 06/29/2016 16:05

2.1.1.3 Sample Data

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12817.D Vial: 20
Acq On : 23 Jun 2016 1:14 Operator: JDS
Sample : L16061026-01 A 826-LOW Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58:43 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	496621	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	442694	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	262492	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	147182	27.0799 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 108.32%		
43) 1,2-Dichloroethane-d4	10.24	65	161862	26.1184 ug/L	0.01
Spiked Amount 25.000	Range 80 - 120		Recovery = 104.48%		
57) Toluene-d8	12.47	98	483174	25.1194 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 100.48%		
78) p-Bromofluorobenzene	15.64	95	184611	22.8152 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 91.28%		

Target Compounds

					Qvalue
3) Chloromethane	3.72	50	3770	0.6050 ug/L	92
4) Vinyl Chloride	3.97	62	856	0.1654 ug/L #	44
13) Acetone	6.34	43	478	0.4053 ug/L #	49
46) Trichloroethene	11.09	130	1544	0.2491 ug/L	86

(#) = qualifier out of range (m) = manual integration
11M12817.D 8260WT.M Fri Jun 24 14:58:44 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12817.D Vial: 20
Acq On : 23 Jun 2016 1:14 Operator: JDS
Sample : L16061026-01 A 826-LOW Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration





#3
Chloromethane
Concen: 0.60 ug/L
RT: 3.72 min Scan# 111
Delta R.T. 0.00 min
Lab File: 11M12817.D
Acq: 23 Jun 2016 1:14

Tgt Ion: 50 Resp: 3770

Ion	Ratio	Lower	Upper
50	100		
52	38.0	20.0	46.6





#4
Vinyl Chloride
Concen: 0.17 ug/L
RT: 3.97 min Scan# 135
Delta R.T. 0.01 min
Lab File: 11M12817.D
Acq: 23 Jun 2016 1:14

Tgt Ion: 62 Resp: 856

Ion	Ratio	Lower	Upper
62	100		
64	0.0	18.4	43.0#





#13
Acetone
Concen: 0.41 ug/L
RT: 6.34 min Scan# 364
Delta R.T. 0.00 min
Lab File: 11M12817.D
Acq: 23 Jun 2016 1:14

Tgt Ion: 43 Resp: 478

Ion	Ratio	Lower	Upper
43	100		
58	0.0	15.8	36.8#





#46
Trichloroethene
Concen: 0.25 ug/L
RT: 11.09 min Scan# 824
Delta R.T. 0.00 min
Lab File: 11M12817.D
Acq: 23 Jun 2016 1:14

Tgt Ion:130 Resp: 1544

Ion	Ratio	Lower	Upper
130	100		
132	114.9	58.3	136.1
95	101.5	55.6	129.8



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12813.D Vial: 16
Acq On : 22 Jun 2016 23:06 Operator: JDS
Sample : L16060974-03 A 826-SPE Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58:33 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	555060	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	484683	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	286068	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	158938	26.1641 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 104.64%		
43) 1,2-Dichloroethane-d4	10.24	65	170923	24.6767 ug/L	0.01
Spiked Amount 25.000	Range 80 - 120		Recovery = 98.72%		
57) Toluene-d8	12.47	98	531603	25.2429 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 100.96%		
78) p-Bromofluorobenzene	15.64	95	200718	22.7614 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 91.04%		

Target Compounds					Qvalue
13) Acetone	6.33	43	793	0.6017 ug/L #	49

(#) = qualifier out of range (m) = manual integration
11M12813.D 8260WT.M Fri Jun 24 14:58:34 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12813.D Vial: 16
Acq On : 22 Jun 2016 23:06 Operator: JDS
Sample : L16060974-03 A 826-SPE Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration



11M12813.D 8260WT.M Fri Jun 24 14:58:34 2016 Page 2



#13
Acetone
Concen: 0.60 ug/L
RT: 6.33 min Scan# 363
Delta R.T. -0.01 min
Lab File: 11M12813.D
Acq: 22 Jun 2016 23:06

Tgt Ion: 43 Resp: 793

Ion	Ratio	Lower	Upper
43	100		
58	0.0	15.8	36.8#



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12815.D Vial: 18
Acq On : 23 Jun 2016 00:10 Operator: JDS
Sample : L16061026-04 A 5X 826-LOW Inst : hpms11
Misc : 1,5 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58:38 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	534594	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	471668	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	276226	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	155657	26.6049 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 106.40%		
43) 1,2-Dichloroethane-d4	10.23	65	167579	25.1202 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 100.48%		
57) Toluene-d8	12.47	98	525050	25.6197 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 102.48%		
78) p-Bromofluorobenzene	15.64	95	201645	23.6813 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 94.72%		

Target Compounds

					Qvalue
12) 1,1,2-Trichloro-1,2,2-Trif	6.24	101	36649	6.8020 ug/L	95
13) Acetone	6.36	43	189	0.1489 ug/L #	49
14) 1,1-Dichloroethene	6.55	61	1475	0.1507 ug/L #	46
32) cis-1,2-Dichloroethene	9.15	96	5996	1.0040 ug/L	76
33) Chloroform	9.35	83	7004	0.6563 ug/L	96
36) Tetrahydrofuran	9.60	42	973	0.7180 ug/L #	66
44) 1,2-Dichloroethane	10.34	62	44118	5.0896 ug/L	88
46) Trichloroethene	11.09	130	2136212	320.1444 ug/L	98
47) Methylcyclohexane	11.08	83	21923	2.6622 ug/L #	1

(#) = qualifier out of range (m) = manual integration
11M12815.D 8260WT.M Fri Jun 24 14:58:39 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12815.D Vial: 18
Acq On : 23 Jun 2016 00:10 Operator: JDS
Sample : L16061026-04 A 5X 826-LOW Inst : hpms11
Misc : 1,5 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration





#12
1,1,2-Trichloro-1,2,2-Trifluoroethane
Concen: 6.80 ug/L
RT: 6.24 min Scan# 354
Delta R.T. -0.01 min
Lab File: 11M12815.D
Acq: 23 Jun 2016 00:10

Tgt Ion:101 Resp: 36649

Ion	Ratio	Lower	Upper
101	100		
151	95.9	50.7	130.7





#13
Acetone
Concen: 0.15 ug/L
RT: 6.36 min Scan# 366
Delta R.T. 0.02 min
Lab File: 11M12815.D
Acq: 23 Jun 2016 00:10

Tgt Ion: 43 Resp: 189

Ion	Ratio	Lower	Upper
43	100		
58	0.0	15.8	36.8#





#14
1,1-Dichloroethene
Concen: 0.15 ug/L
RT: 6.55 min Scan# 384
Delta R.T. -0.01 min
Lab File: 11M12815.D
Acq: 23 Jun 2016 00:10

Tgt Ion: 61 Resp: 1475

Ion	Ratio	Lower	Upper
61	100		
96	13.2	30.1	70.3#





#32
cis-1,2-Dichloroethene
Concen: 1.00 ug/L
RT: 9.15 min Scan# 636
Delta R.T. 0.00 min
Lab File: 11M12815.D
Acq: 23 Jun 2016 00:10

Tgt Ion: 96 Resp: 5996

Ion	Ratio	Lower	Upper
96	100		
61	142.2	105.2	245.6





#33
Chloroform
Concen: 0.66 ug/L
RT: 9.35 min Scan# 655
Delta R.T. 0.00 min
Lab File: 11M12815.D
Acq: 23 Jun 2016 00:10

Tgt Ion: 83 Resp: 7004

Ion	Ratio	Lower	Upper
83	100		
85	62.3	39.2	91.4





#36
Tetrahydrofuran
Concen: 0.72 ug/L
RT: 9.60 min Scan# 679
Delta R.T. 0.00 min
Lab File: 11M12815.D
Acq: 23 Jun 2016 00:10

Tgt Ion: 42 Resp: 973

Ion	Ratio	Lower	Upper
42	100		
71	0.0	16.4	38.2#
72	19.5	16.6	38.8





#44
1,2-Dichloroethane
Concen: 5.09 ug/L
RT: 10.34 min Scan# 751
Delta R.T. 0.00 min
Lab File: 11M12815.D
Acq: 23 Jun 2016 00:10

Tgt Ion: 62 Resp: 44118

Ion	Ratio	Lower	Upper
62	100		
64	29.1	18.7	43.7
49	26.3	22.3	52.1





#46
Trichloroethene
Concen: 320.14 ug/L
RT: 11.09 min Scan# 824
Delta R.T. 0.00 min
Lab File: 11M12815.D
Acq: 23 Jun 2016 00:10

Tgt Ion:130 Resp: 2136212

Ion	Ratio	Lower	Upper
130	100		
132	96.8	58.3	136.1
95	89.6	55.6	129.8





#47
Methylcyclohexane
Concen: 2.66 ug/L
RT: 11.08 min Scan# 823
Delta R.T. -0.08 min
Lab File: 11M12815.D
Acq: 23 Jun 2016 00:10

Tgt Ion: 83 Resp: 21923

Ion	Ratio	Lower	Upper
83	100		
55	0.0	66.5	155.1#
98	236.9	29.9	69.7#



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062316\11M12835.D Vial: 9
Acq On : 23 Jun 2016 19:11 Operator: JDS
Sample : L16061026-04 B D1 20X 826-LOW Inst : hpms11
Misc : 1,20 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 15:00:33 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	507627	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	457128	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	268568	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	148019	26.6434 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 106.56%		
43) 1,2-Dichloroethane-d4	10.24	65	163290	25.7776 ug/L	0.01
Spiked Amount 25.000	Range 80 - 120		Recovery = 103.12%		
57) Toluene-d8	12.47	98	498634	25.1047 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 100.40%		
78) p-Bromofluorobenzene	15.64	95	195933	23.6666 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 94.68%		

Target Compounds

					Qvalue
3) Chloromethane	3.74	50	1416	0.2223 ug/L	97
6) Bromomethane	4.86	94	830	0.2282 ug/L	77
12) 1,1,2-Trichloro-1,2,2-Trif	6.25	101	9036	1.7662 ug/L	96
13) Acetone	6.34	43	1216	1.0088 ug/L #	49
15) Tert-Butyl Alcohol	6.64	59	889	2.6663 ug/L #	56
17) Iodomethane	7.07	142	1039	2.4967 ug/L #	29
18) Methyl acetate	7.07	43	472	Below Cal #	71
20) Carbon Disulfide	7.36	76	2264	0.1455 ug/L #	75
29) 2-Butanone	8.87	43	280	0.1537 ug/L #	62
30) Propionitrile	8.99	54	676	1.2666 ug/L #	60
32) cis-1,2-Dichloroethene	9.15	96	1534	0.2705 ug/L	88
33) Chloroform	9.36	83	1713	0.1690 ug/L	71
36) Tetrahydrofuran	9.61	42	1610	1.2512 ug/L #	47
42) Tert-Amyl-Methyl ether	10.14	73	1878	0.1553 ug/L #	63
44) 1,2-Dichloroethane	10.34	62	10039	1.2197 ug/L	86
46) Trichloroethene	11.09	130	515068	81.2916 ug/L	97
47) Methylcyclohexane	11.08	83	4917	0.6288 ug/L #	1
62) 2-Hexanone	12.86	43	403	0.1337 ug/L #	34
64) Tetrachloroethene	13.34	164	809	0.1503 ug/L #	53
96) 1,2,4-Trichlorobenzene	19.54	180	1319	0.1255 ug/L #	43
98) Naphthalene	19.89	128	4993	0.2560 ug/L #	77
99) 1,2,3-Trichlorobenzene	20.18	180	1619	0.1679 ug/L #	63

(#) = qualifier out of range (m) = manual integration
11M12835.D 8260WT.M Fri Jun 24 15:00:35 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062316\11M12835.D Vial: 9
Acq On : 23 Jun 2016 19:11 Operator: JDS
Sample : L16061026-04 B D1 20X 826-LOW Inst : hpms11
Misc : 1,20 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 15:00 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration





#3
Chloromethane
Concen: 0.22 ug/L
RT: 3.74 min Scan# 113
Delta R.T. 0.02 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 50 Resp: 1416

Ion	Ratio	Lower	Upper
50	100		
52	31.5	20.0	46.6





#6
Bromomethane
Concen: 0.23 ug/L
RT: 4.86 min Scan# 221
Delta R.T. 0.01 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 94 Resp: 830

Ion	Ratio	Lower	Upper
94	100		
96	71.9	56.5	131.9





#12
1,1,2-Trichloro-1,2,2-Trifluoroethane
Concen: 1.77 ug/L
RT: 6.25 min Scan# 355
Delta R.T. 0.00 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion:101 Resp: 9036

Ion	Ratio	Lower	Upper
101	100		
151	94.6	50.7	130.7





#13
Acetone
Concen: 1.01 ug/L
RT: 6.34 min Scan# 364
Delta R.T. 0.00 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 43 Resp: 1216

Ion	Ratio	Lower	Upper
43	100		
58	0.0	15.8	36.8#





#15
Tert-Butyl Alcohol
Concen: 2.67 ug/L
RT: 6.64 min Scan# 393
Delta R.T. -0.02 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 59 Resp: 889

Ion	Ratio	Lower	Upper
59	100		
41	0.0	15.3	35.7#
57	0.0	6.2	14.4#





#17
Iodomethane
Concen: 2.50 ug/L
RT: 7.07 min Scan# 435
Delta R.T. 0.01 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion:142 Resp: 1039

Ion	Ratio	Lower	Upper
142	100		
127	0.0	29.0	67.6#





#18
Methyl acetate
Concen: Below Cal
RT: 7.07 min Scan# 435
Delta R.T. 0.01 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 43 Resp: 472

Ion	Ratio	Lower	Upper
43	100		
74	0.0	8.2	19.0#
59	0.0	3.9	9.1#





#20
Carbon Disulfide
Concen: 0.15 ug/L
RT: 7.36 min Scan# 463
Delta R.T. 0.00 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 76 Resp: 2264

Ion	Ratio	Lower	Upper
76	100		
78	0.0	5.5	12.9#





#29
2-Butanone
Concen: 0.15 ug/L
RT: 8.87 min Scan# 609
Delta R.T. 0.00 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 43 Resp: 280

Ion	Ratio	Lower	Upper
43	100		
72	0.0	9.6	22.4#





#30
Propionitrile
Concen: 1.27 ug/L
RT: 8.99 min Scan# 620
Delta R.T. 0.01 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 54 Resp: 676

Ion	Ratio	Lower	Upper
54	100		
55	0.0	10.3	24.1#





#32
cis-1,2-Dichloroethene
Concen: 0.27 ug/L
RT: 9.15 min Scan# 636
Delta R.T. 0.00 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 96 Resp: 1534

Ion	Ratio	Lower	Upper
96	100		
61	158.9	105.2	245.6





#33
Chloroform
Concen: 0.17 ug/L
RT: 9.36 min Scan# 656
Delta R.T. 0.01 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 83 Resp: 1713

Ion	Ratio	Lower	Upper
83	100		
85	42.5	39.2	91.4





#36
Tetrahydrofuran
Concen: 1.25 ug/L
RT: 9.61 min Scan# 680
Delta R.T. 0.01 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 42 Resp: 1610

Ion	Ratio	Lower	Upper
42	100		
71	0.0	16.4	38.2#
72	0.0	16.6	38.8#





#42
Tert-Amyl-Methyl ether
Concen: 0.16 ug/L
RT: 10.14 min Scan# 732
Delta R.T. 0.01 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 73 Resp: 1878

Ion	Ratio	Lower	Upper
73	100		
55	11.1	20.1	46.9#
57	0.0	1.4	3.4#





#44
1,2-Dichloroethane
Concen: 1.22 ug/L
RT: 10.34 min Scan# 751
Delta R.T. 0.00 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 62 Resp: 10039

Ion	Ratio	Lower	Upper
62	100		
64	35.4	18.7	43.7
49	25.2	22.3	52.1





#46
Trichloroethene
Concen: 81.29 ug/L
RT: 11.09 min Scan# 824
Delta R.T. 0.00 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion:130 Resp: 515068

Ion	Ratio	Lower	Upper
130	100		
132	96.6	58.3	136.1
95	87.9	55.6	129.8





#47
Methylcyclohexane
Concen: 0.63 ug/L
RT: 11.08 min Scan# 823
Delta R.T. -0.08 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 83 Resp: 4917

Ion	Ratio	Lower	Upper
83	100		
55	0.0	66.5	155.1#
98	254.8	29.9	69.7#





#62
2-Hexanone
Concen: 0.13 ug/L
RT: 12.86 min Scan# 995
Delta R.T. -0.00 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion: 43 Resp: 403

Ion	Ratio	Lower	Upper
43	100		
58	0.0	29.0	67.6#
85	0.0	3.2	7.6#





#64
Tetrachloroethene
Concen: 0.15 ug/L
RT: 13.34 min Scan# 1041
Delta R.T. 0.00 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion:164 Resp: 809

Ion	Ratio	Lower	Upper
164	100		
129	142.4	57.9	135.1#





#96
1,2,4-Trichlorobenzene
Concen: 0.13 ug/L
RT: 19.54 min Scan# 1641
Delta R.T. -0.01 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion:180 Resp: 1319

Ion	Ratio	Lower	Upper
180	100		
145	0.0	19.1	44.7#





#98
Naphthalene
Concen: 0.26 ug/L
RT: 19.89 min Scan# 1675
Delta R.T. -0.01 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion:128 Resp: 4993

Ion	Ratio	Lower	Upper
128	100		
127	3.7	11.7	14.3#





#99
1,2,3-Trichlorobenzene
Concen: 0.17 ug/L
RT: 20.18 min Scan# 1703
Delta R.T. -0.01 min
Lab File: 11M12835.D
Acq: 23 Jun 2016 19:11

Tgt Ion:180 Resp: 1619

Ion	Ratio	Lower	Upper
180	100		
145	12.5	20.3	47.3#



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12816.D Vial: 19
Acq On : 23 Jun 2016 00:42 Operator: JDS
Sample : L16061026-05 A 100X 826-LOW Inst : hpms11
Misc : 1,100 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58:40 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	517460	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	458713	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	269312	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	150082	26.5014 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 106.00%		
43) 1,2-Dichloroethane-d4	10.24	65	165759	25.6701 ug/L	0.01
Spiked Amount 25.000	Range 80 - 120		Recovery = 102.68%		
57) Toluene-d8	12.47	98	507453	25.4604 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 101.84%		
78) p-Bromofluorobenzene	15.64	95	195293	23.5241 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 94.08%		

Target Compounds

					Qvalue
4) Vinyl Chloride	3.96	62	4784	0.8872 ug/L #	77
12) 1,1,2-Trichloro-1,2,2-Trif	6.25	101	2825	0.5417 ug/L	94
14) 1,1-Dichloroethene	6.56	61	12701	1.3408 ug/L	99
23) trans-1,2-Dichloroethene	7.74	96	796	0.1518 ug/L	92
27) 1,1-Dichloroethane	8.34	63	2267	0.2155 ug/L #	52
32) cis-1,2-Dichloroethene	9.15	96	87564	15.1476 ug/L	80
33) Chloroform	9.35	83	3921	0.3796 ug/L	96
36) Tetrahydrofuran	9.60	42	1032	0.7867 ug/L #	47
44) 1,2-Dichloroethane	10.34	62	2692	0.3208 ug/L #	76
46) Trichloroethene	11.09	130	1107013	171.3964 ug/L	98
47) Methylcyclohexane	11.09	83	11347	1.4235 ug/L #	1

(#) = qualifier out of range (m) = manual integration
11M12816.D 8260WT.M Fri Jun 24 14:58:41 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12816.D Vial: 19
Acq On : 23 Jun 2016 00:42 Operator: JDS
Sample : L16061026-05 A 100X 826-LOW Inst : hpms11
Misc : 1,100 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration





#4
Vinyl Chloride
Concen: 0.89 ug/L
RT: 3.96 min Scan# 134
Delta R.T. 0.00 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion: 62 Resp: 4784

Ion	Ratio	Lower	Upper
62	100		
64	18.1	18.4	43.0#





#12
1,1,2-Trichloro-1,2,2-Trifluoroethane
Concen: 0.54 ug/L
RT: 6.25 min Scan# 355
Delta R.T. 0.00 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion:101 Resp: 2825

Ion	Ratio	Lower	Upper
101	100		
151	96.1	50.7	130.7





#14
1,1-Dichloroethene
Concen: 1.34 ug/L
RT: 6.56 min Scan# 385
Delta R.T. 0.00 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion: 61 Resp: 12701

Ion	Ratio	Lower	Upper
61	100		
96	49.6	30.1	70.3





#23
trans-1,2-Dichloroethene
Concen: 0.15 ug/L
RT: 7.74 min Scan# 500
Delta R.T. 0.00 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion: 96 Resp: 796

Ion	Ratio	Lower	Upper
96	100		
61	178.0	100.6	234.6





#27
1,1-Dichloroethane
Concen: 0.22 ug/L
RT: 8.34 min Scan# 558
Delta R.T. 0.00 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion: 63 Resp: 2267

Ion	Ratio	Lower	Upper
63	100		
65	0.0	17.8	41.6#
83	0.0	7.4	17.4#





#32
cis-1,2-Dichloroethene
Concen: 15.15 ug/L
RT: 9.15 min Scan# 636
Delta R.T. 0.00 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion: 96 Resp: 87564

Ion	Ratio	Lower	Upper
96	100		
61	146.9	105.2	245.6





#33
Chloroform
Concen: 0.38 ug/L
RT: 9.35 min Scan# 655
Delta R.T. 0.00 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion: 83 Resp: 3921

Ion	Ratio	Lower	Upper
83	100		
85	68.1	39.2	91.4





#36
Tetrahydrofuran
Concen: 0.79 ug/L
RT: 9.60 min Scan# 679
Delta R.T. 0.00 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion: 42 Resp: 1032

Ion	Ratio	Lower	Upper
42	100		
71	0.0	16.4	38.2#
72	0.0	16.6	38.8#





#44
1,2-Dichloroethane
Concen: 0.32 ug/L
RT: 10.34 min Scan# 751
Delta R.T. 0.00 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion: 62 Resp: 2692

Ion	Ratio	Lower	Upper
62	100		
64	24.6	18.7	43.7
49	16.6	22.3	52.1#





#46
Trichloroethene
Concen: 171.40 ug/L
RT: 11.09 min Scan# 824
Delta R.T. 0.00 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion:130 Resp: 1107013

Ion	Ratio	Lower	Upper
130	100		
132	98.0	58.3	136.1
95	89.6	55.6	129.8





#47
Methylcyclohexane
Concen: 1.42 ug/L
RT: 11.09 min Scan# 824
Delta R.T. -0.07 min
Lab File: 11M12816.D
Acq: 23 Jun 2016 00:42

Tgt Ion: 83 Resp: 11347

Ion	Ratio	Lower	Upper
83	100		
55	0.0	66.5	155.1#
98	238.3	29.9	69.7#



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12814.D Vial: 17
Acq On : 22 Jun 2016 23:38 Operator: JDS
Sample : L16060974-07 A 826-SPE Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58:35 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	557279	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	492737	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	285222	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	161635	26.5021 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 106.00%		
43) 1,2-Dichloroethane-d4	10.23	65	172830	24.8527 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 99.40%		
57) Toluene-d8	12.47	98	536215	25.0457 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 100.20%		
78) p-Bromofluorobenzene	15.64	95	205601	23.3843 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 93.52%		

Target Compounds

					Qvalue
3) Chloromethane	3.73	50	925	0.1323 ug/L #	41
12) 1,1,2-Trichloro-1,2,2-Trif	6.26	101	9847	1.7532 ug/L	98
13) Acetone	6.36	43	380	0.2872 ug/L #	49
14) 1,1-Dichloroethene	6.56	61	1518	0.1488 ug/L #	27
32) cis-1,2-Dichloroethene	9.15	96	89484	14.3737 ug/L	75
44) 1,2-Dichloroethane	10.34	62	10814	1.1968 ug/L	92
46) Trichloroethene	11.09	130	556994	80.0762 ug/L	98
47) Methylcyclohexane	11.09	83	5328	0.6207 ug/L #	1

(#) = qualifier out of range (m) = manual integration
11M12814.D 8260WT.M Fri Jun 24 14:58:36 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12814.D Vial: 17
Acq On : 22 Jun 2016 23:38 Operator: JDS
Sample : L16060974-07 A 826-SPE Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration





#3
Chloromethane
Concen: 0.13 ug/L
RT: 3.73 min Scan# 112
Delta R.T. 0.01 min
Lab File: 11M12814.D
Acq: 22 Jun 2016 23:38

Tgt Ion: 50 Resp: 925

Ion	Ratio	Lower	Upper
50	100		
52	0.0	20.0	46.6#





#12
1,1,2-Trichloro-1,2,2-Trifluoroethane
Concen: 1.75 ug/L
RT: 6.26 min Scan# 356
Delta R.T. 0.01 min
Lab File: 11M12814.D
Acq: 22 Jun 2016 23:38

Tgt Ion:101 Resp: 9847

Ion	Ratio	Lower	Upper
101	100		
151	89.0	50.7	130.7





#13
Acetone
Concen: 0.29 ug/L
RT: 6.36 min Scan# 366
Delta R.T. 0.02 min
Lab File: 11M12814.D
Acq: 22 Jun 2016 23:38

Tgt Ion: 43 Resp: 380

Ion	Ratio	Lower	Upper
43	100		
58	0.0	15.8	36.8#





#14
1,1-Dichloroethene
Concen: 0.15 ug/L
RT: 6.56 min Scan# 385
Delta R.T. 0.00 min
Lab File: 11M12814.D
Acq: 22 Jun 2016 23:38

Tgt Ion: 61 Resp: 1518

Ion	Ratio	Lower	Upper
61	100		
96	0.0	30.1	70.3#





#32
cis-1,2-Dichloroethene
Concen: 14.37 ug/L
RT: 9.15 min Scan# 636
Delta R.T. 0.00 min
Lab File: 11M12814.D
Acq: 22 Jun 2016 23:38

Tgt Ion: 96 Resp: 89484

Ion	Ratio	Lower	Upper
96	100		
61	141.0	105.2	245.6





#44
1,2-Dichloroethane
Concen: 1.20 ug/L
RT: 10.34 min Scan# 751
Delta R.T. 0.00 min
Lab File: 11M12814.D
Acq: 22 Jun 2016 23:38

Tgt Ion: 62 Resp: 10814

Ion	Ratio	Lower	Upper
62	100		
64	31.0	18.7	43.7
49	28.3	22.3	52.1





#46
Trichloroethene
Concen: 80.08 ug/L
RT: 11.09 min Scan# 824
Delta R.T. 0.00 min
Lab File: 11M12814.D
Acq: 22 Jun 2016 23:38

Tgt Ion:130 Resp: 556994

Ion	Ratio	Lower	Upper
130	100		
132	96.5	58.3	136.1
95	89.1	55.6	129.8





#47
Methylcyclohexane
Concen: 0.62 ug/L
RT: 11.09 min Scan# 824
Delta R.T. -0.07 min
Lab File: 11M12814.D
Acq: 22 Jun 2016 23:38

Tgt Ion: 83 Resp: 5328

Ion	Ratio	Lower	Upper
83	100		
55	0.0	66.5	155.1#
98	246.2	29.9	69.7#



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12808.D Vial: 11
Acq On : 22 Jun 2016 20:27 Operator: JDS
Sample : L16061026-08 A TB 826-LOW Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58:21 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	611363	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	533935	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	313293	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	171569	25.6423 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 102.56%		
43) 1,2-Dichloroethane-d4	10.23	65	176032	23.0738 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 92.28%		
57) Toluene-d8	12.47	98	587180	25.3101 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 101.24%		
78) p-Bromofluorobenzene	15.64	95	224618	23.2582 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 93.04%		

Target Compounds

					Qvalue
3) Chloromethane	3.72	50	1080	0.1408 ug/L #	1
6) Bromomethane	4.86	94	701	0.1600 ug/L #	31
13) Acetone	6.35	43	697	0.4801 ug/L #	49
16) Dimethyl Sulfide	6.81	62	1450	0.3244 ug/L #	1
17) Iodomethane	7.07	142	414	2.3406 ug/L #	29
36) Tetrahydrofuran	9.60	42	2429	1.5673 ug/L	85
58) Toluene	12.56	91	4943	0.1829 ug/L	98
98) Naphthalene	19.89	128	3247	0.1427 ug/L #	67

(#) = qualifier out of range (m) = manual integration
11M12808.D 8260WT.M Fri Jun 24 14:58:22 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12808.D Vial: 11
Acq On : 22 Jun 2016 20:27 Operator: JDS
Sample : L16061026-08 A TB 826-LOW Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration





#3
Chloromethane
Concen: 0.14 ug/L
RT: 3.72 min Scan# 111
Delta R.T. 0.00 min
Lab File: 11M12808.D
Acq: 22 Jun 2016 20:27

Tgt Ion: 50 Resp: 1080

Ion	Ratio	Lower	Upper
50	100		
52	113.8	20.0	46.6#





#6
Bromomethane
Concen: 0.16 ug/L
RT: 4.86 min Scan# 221
Delta R.T. 0.01 min
Lab File: 11M12808.D
Acq: 22 Jun 2016 20:27

Tgt Ion: 94 Resp: 701

Ion	Ratio	Lower	Upper
94	100		
96	27.1	56.5	131.9#





#13
Acetone
Concen: 0.48 ug/L
RT: 6.35 min Scan# 365
Delta R.T. 0.01 min
Lab File: 11M12808.D
Acq: 22 Jun 2016 20:27

Tgt Ion: 43 Resp: 697

Ion	Ratio	Lower	Upper
43	100		
58	0.0	15.8	36.8#





#16
Dimethyl Sulfide
Concen: 0.32 ug/L
RT: 6.81 min Scan# 410
Delta R.T. 0.00 min
Lab File: 11M12808.D
Acq: 22 Jun 2016 20:27

Tgt Ion: 62 Resp: 1450

Ion	Ratio	Lower	Upper
62	100		
47	0.0	73.4	171.2#





#17
Iodomethane
Concen: 2.34 ug/L
RT: 7.07 min Scan# 435
Delta R.T. 0.01 min
Lab File: 11M12808.D
Acq: 22 Jun 2016 20:27

Tgt Ion:142 Resp: 414

Ion	Ratio	Lower	Upper
142	100		
127	0.0	29.0	67.6#





#36
Tetrahydrofuran
Concen: 1.57 ug/L
RT: 9.60 min Scan# 679
Delta R.T. 0.00 min
Lab File: 11M12808.D
Acq: 22 Jun 2016 20:27

Tgt Ion: 42 Resp: 2429

Ion	Ratio	Lower	Upper
42	100		
71	17.3	16.4	38.2
72	22.3	16.6	38.8





#58
Toluene
Concen: 0.18 ug/L
RT: 12.56 min Scan# 966
Delta R.T. 0.00 min
Lab File: 11M12808.D
Acq: 22 Jun 2016 20:27

Tgt Ion: 91 Resp: 4943

Ion	Ratio	Lower	Upper
91	100		
92	61.3	35.7	83.3





#98
Naphthalene
Concen: 0.14 ug/L
RT: 19.89 min Scan# 1675
Delta R.T. -0.01 min
Lab File: 11M12808.D
Acq: 22 Jun 2016 20:27

Tgt Ion:128 Resp: 3247

Ion	Ratio	Lower	Upper
128	100		
127	0.0	11.7	14.3#



2.1.1.4 Standards Data

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11837.D Vial: 3
Acq On : 13 May 2016 14:40 Operator: JDS
Sample : WG568769-02 0.3ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04:20 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	538041	25.00 ug/L	-0.01
56) Chlorobenzene-d5	14.25	117	450050	25.00 ug/L	-0.01
76) 1,4-Dichlorobenzene-d4	17.06	152	251600	25.00 ug/L	-0.01

System Monitoring Compounds					
37) Dibromofluoromethane	0.00	111	0	0.0000 ug/L	
Spiked Amount 25.000	Range 86 - 118		Recovery =	0.00%#	
43) 1,2-Dichloroethane-d4	0.00	65	0	0.0000 ug/L	
Spiked Amount 25.000	Range 80 - 120		Recovery =	0.00%#	
57) Toluene-d8	12.47	98	210	0.0099 ug/L	-0.01
Spiked Amount 25.000	Range 88 - 110		Recovery =	0.04%#	
78) p-Bromofluorobenzene	15.64	95	566	0.0689 ug/L	-0.01
Spiked Amount 25.000	Range 86 - 115		Recovery =	0.28%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	2475	0.3135 ug/L #	66
3) Chloromethane	3.72	50	4639	0.6471 ug/L	91
4) Vinyl Chloride	3.96	62	1882	0.3288 ug/L #	44
5) 1,3-Butadiene	4.01	54	1564	0.2900 ug/L #	19
6) Bromomethane	4.84	94	1717	0.4723 ug/L	83
7) Chloroethane	5.00	64	616	0.1589 ug/L #	45
8) Trichlorofluoromethane	5.49	101	3531	0.3330 ug/L	97
10) Isoprene	6.04	67	1746	0.2381 ug/L	92
12) 1,1,2-Trichloro-1,2,2-Trif	6.26	101	1344	0.2515 ug/L	96
13) Acetone	6.34	43	2389	1.4850 ug/L #	45
14) 1,1-Dichloroethene	6.56	61	3780	0.3817 ug/L	78
16) Dimethyl Sulfide	6.80	62	762	0.1697 ug/L	78
18) Methyl acetate	7.02	43	24685	4.4975 ug/L #	70
19) Methylene Chloride	7.31	84	1766	0.3184 ug/L	74
20) Carbon Disulfide	7.36	76	5848	0.3373 ug/L	92
22) Methyl Tert Butyl Ether	7.54	73	3068	0.2131 ug/L #	48
23) trans-1,2-Dichloroethene	7.75	96	1449	0.2556 ug/L	91
24) n-Hexane	7.82	57	3745	0.4031 ug/L #	73
26) Vinyl Acetate	8.30	43	2410	0.6995 ug/L #	78
27) 1,1-Dichloroethane	8.35	63	3312	0.2969 ug/L #	76
31) 2,2-Dichloropropane	9.10	77	2449	0.3040 ug/L #	42
32) cis-1,2-Dichloroethene	9.15	96	1729	0.2748 ug/L	96
33) Chloroform	9.35	83	3792	0.3639 ug/L	92
35) Bromochloromethane	9.56	130	948	0.2355 ug/L #	68
38) 1,1,1-Trichloroethane	9.85	97	3449	0.3423 ug/L #	79
39) Cyclohexane	9.88	56	3639	0.2986 ug/L #	84
40) 1,1-Dichloropropene	10.04	75	2365	0.3106 ug/L #	53
41) Carbon Tetrachloride	10.17	117	2613	0.2748 ug/L	87
44) 1,2-Dichloroethane	10.35	62	2493	0.2788 ug/L #	73
45) Benzene	10.38	78	7133	0.3275 ug/L	99
46) Trichloroethene	11.09	130	1973	0.2780 ug/L	93
47) Methylcyclohexane	11.17	83	2570	0.2986 ug/L	84
48) 1,2-Dichloropropane	11.29	63	1665	0.2692 ug/L	87
50) Bromodichloromethane	11.57	83	2418	0.2950 ug/L #	82
51) Dibromomethane	11.65	93	756	0.2331 ug/L	88
54) cis-1,3-Dichloropropene	12.17	75	1869	0.2166 ug/L #	82
55) Dimethyl Disulfide	12.42	79	1028	0.1910 ug/L	95
58) Toluene	12.57	91	7355	0.3156 ug/L	91
59) Ethyl Methacrylate	12.64	69	1041	0.1887 ug/L	88
60) trans-1,3-Dichloropropene	12.72	75	1951	0.2556 ug/L #	51

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11837.D Vial: 3
Acq On : 13 May 2016 14:40 Operator: JDS
Sample : WG568769-02 0.3ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04:20 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
61)	1,1,2-Trichloroethane	12.92	97	1087	0.2492 ug/L		87
62)	2-Hexanone	12.86	43	687	0.1910 ug/L	#	29
63)	1,3-Dichloropropane	13.21	76	1673	0.2309 ug/L		94
64)	Tetrachloroethene	13.34	164	1703	0.3238 ug/L		94
65)	Dibromochloromethane	13.59	129	1515	0.2406 ug/L		92
66)	1,2-Dibromoethane	13.82	107	966	0.2181 ug/L		98
67)	1-Chlorohexane	13.89	91	2128	0.2784 ug/L		97
68)	Chlorobenzene	14.29	112	5248	0.3062 ug/L		99
69)	1,1,1,2-Tetrachloroethane	14.32	131	2025	0.3082 ug/L		98
70)	Ethylbenzene	14.31	106	2602	0.2992 ug/L		88
71)	m-,p-Xylene	14.40	106	6296	0.6083 ug/L		82
72)	o-Xylene	14.92	106	3032	0.2937 ug/L		95
73)	Styrene	14.95	104	4953	0.2828 ug/L		95
75)	Isopropylbenzene	15.31	105	7848	0.3006 ug/L		93
77)	1,1,2,2-Tetrachloroethane	15.52	83	1145	0.2759 ug/L		68
81)	n-Propylbenzene	15.79	91	9421	0.3332 ug/L		94
82)	Bromobenzene	15.92	156	2002	0.2551 ug/L		66
83)	1,3,5-Trimethylbenzene	15.95	105	6742	0.3138 ug/L		91
84)	2-Chlorotoluene	16.05	91	6735	0.3228 ug/L		96
85)	4-Chlorotoluene	16.09	91	6023	0.3518 ug/L		99
86)	a-Methylstyrene	16.34	118	2850	0.2387 ug/L		95
87)	tert-Butylbenzene	16.39	134	1197	0.2539 ug/L		64
88)	1,2,4-Trimethylbenzene	16.44	105	6992	0.3180 ug/L		88
89)	sec-Butylbenzene	16.65	105	8335	0.3278 ug/L		97
90)	p-Isopropyltoluene	16.79	119	7094	0.3065 ug/L		96
91)	1,3-Dichlorobenzene	16.98	146	4505	0.3090 ug/L		96
92)	1,4-Dichlorobenzene	17.10	146	4475	0.3006 ug/L	#	1
93)	n-Butylbenzene	17.28	91	7076	0.3396 ug/L	#	93
94)	1,2-Dichlorobenzene	17.56	146	4016	0.2922 ug/L		100
96)	1,2,4-Trichlorobenzene	19.55	180	2964	0.2902 ug/L		99
97)	Hexachlorobutadiene	19.69	225	1160	0.2796 ug/L	#	71
98)	Naphthalene	19.90	128	4679	0.2394 ug/L	#	93
99)	1,2,3-Trichlorobenzene	20.19	180	2578	0.2696 ug/L		87

(#) = qualifier out of range (m) = manual integration
11M11837.D 8260WT.M Fri May 13 17:04:21 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11837.D Vial: 3
Acq On : 13 May 2016 14:40 Operator: JDS
Sample : WG568769-02 0.3ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\051316\11M11837.D Vial: 3
Acq On : 13 May 2016 14:40 Operator: JDS
Sample : WG568769-02 0.3ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:40:10 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	-1.0000	0.2936	0.0	0	0.00
3	P	Chloromethane	-1.0000	0.6871	0.0	0	0.00
4	C	Vinyl Chloride	-1.0000	0.3357	0.0	0	0.00
5	T	1,3-Butadiene	-1.0000	0.2805	0.0	0	0.01
6	T	Bromomethane	-1.0000	0.4454	0.0	0	-0.01
7	T	Chloroethane	-1.0000	0.1710	0.0	0	0.00
8	T	Trichlorofluoromethane	-1.0000	0.3160	0.0	0	0.01
9	T	Diethyl ether	-1.0000	0.0000	0.0	0	-6.01#
10	T	Isoprene	-1.0000	0.2535	0.0	0	0.00
11	T	Acrolein	-1.0000	0.0000	0.0	0	-6.23#
12	T	1,1,2-Trichloro-1,2,2-Trifl	-1.0000	0.2478	0.0	0	0.01
13	T	Acetone	-1.0000	1.8699	0.0	0	0.00
14	C	1,1-Dichloroethene	-1.0000	0.3838	0.0	0	0.00
15	T	Tert-Butyl Alcohol	-1.0000	0.0000	0.0	0	-6.66#
16	T	Dimethyl Sulfide	-1.0000	0.1937	0.0	0	-0.01
17	T	Iodomethane	-1.0000	0.0000	0.0	0	-7.06#
18	T	Methyl acetate	-1.0000	5.1442	0.0	0	-0.04
19	T	Methylene Chloride	-1.0000	0.3464	0.0	0	0.00
20	T	Carbon Disulfide	-1.0000	0.3546	0.0	0	0.00
21	T	Acrylonitrile	-1.0000	0.0000	0.0	0	-7.49#
22	T	Methyl Tert Butyl Ether	-1.0000	0.2413	0.0	0	0.02
23	T	trans-1,2-Dichloroethene	-1.0000	0.2658	0.0	0	0.01
24	T	n-Hexane	-1.0000	0.4125	0.0	0	0.00
25	T	Diisopropyl ether	-1.0000	0.0083	0.0	0	0.08
26	T	Vinyl Acetate	-1.0000	2.6933	0.0	0	-0.01
27	P	1,1-Dichloroethane	-1.0000	0.3028	0.0	0	0.01
28	T	Ethyl-Tert-Butyl ether	-1.0000	0.0000	0.0	0	-8.70#
29	T	2-Butanone	-1.0000	0.0000	0.0	0	-8.87#
30	T	Propionitrile	-1.0000	0.0000	0.0	0	-8.97#
31	T	2,2-Dichloropropane	-1.0000	0.2862	0.0	0	0.01
32	T	cis-1,2-Dichloroethene	-1.0000	0.2877	0.0	0	0.00
33	C	Chloroform	0.3000	0.3530	-17.7	100	0.00
34	T	1-Bromopropane	-1.0000	0.0000	0.0	0	-9.48#
35	T	Bromochloromethane	-1.0000	0.2656	0.0	0	-0.01
36	T	Tetrahydrofuran	-1.0000	0.0000	0.0	0	-9.60#
37	S	Dibromofluoromethane	-1.0000	0.0000	0.0	0	-9.63#
38	T	1,1,1-Trichloroethane	-1.0000	0.3263	0.0	0	0.00
39	T	Cyclohexane	-1.0000	0.3159	0.0	0	0.00
40	T	1,1-Dichloropropene	-1.0000	0.3134	0.0	0	0.00
41	T	Carbon Tetrachloride	-1.0000	0.2532	0.0	0	-0.01
42	T	Tert-Amyl-Methyl ether	-1.0000	0.0152	0.0	0	-0.09
43	S	1,2-Dichloroethane-d4	-1.0000	0.0000	0.0	0	-10.23#
44	T	1,2-Dichloroethane	-1.0000	0.2858	0.0	0	0.01
45	T	Benzene	-1.0000	0.3452	0.0	0	0.00
46	T	Trichloroethene	-1.0000	0.2938	0.0	0	0.00
47	T	Methylcyclohexane	-1.0000	0.3101	0.0	0	0.00
48	C	1,2-Dichloropropane	-1.0000	0.2911	0.0	0	0.00
49	T	1,4-Dioxane	-1.0000	0.0000	0.0	0	-11.55#
50	T	Bromodichloromethane	-1.0000	0.3012	0.0	0	0.00
51	T	Dibromomethane	-1.0000	0.2537	0.0	0	0.00
52	T	2-Chloroethyl Vinyl Ether	-1.0000	0.0000	0.0	0	-11.84#
53	T	4-Methyl-2-Pentanone	-1.0000	0.0000	0.0	0	-11.87#
54	T	cis-1,3-Dichloropropene	-1.0000	0.2327	0.0	0	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\051316\11M11837.D Vial: 3
Acq On : 13 May 2016 14:40 Operator: JDS
Sample : WG568769-02 0.3ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:40:10 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	Dimethyl Disulfide	-1.0000	0.2127	0.0	0	0.00
56	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
57	S	Toluene-d8	-1.0000	0.0107	0.0	0	0.00
58	C	Toluene	-1.0000	0.3228	0.0	0	0.01
59	T	Ethyl Methacrylate	-1.0000	0.2140	0.0	0	-0.01
60	T	trans-1,3-Dichloropropene	-1.0000	0.2668	0.0	0	-0.01
61	T	1,1,2-Trichloroethane	-1.0000	0.2691	0.0	0	-0.01
62	T	2-Hexanone	-1.0000	0.2315	0.0	0	0.00
63	T	1,3-Dichloropropane	-1.0000	0.2579	0.0	0	0.00
64	T	Tetrachloroethene	-1.0000	0.3214	0.0	0	0.00
65	T	Dibromochloromethane	-1.0000	0.2459	0.0	0	0.00
66	T	1,2-Dibromoethane	-1.0000	0.2387	0.0	0	0.00
67	T	1-Chlorohexane	-1.0000	0.2833	0.0	0	0.00
68	P	Chlorobenzene	-1.0000	0.3144	0.0	0	0.00
69	T	1,1,1,2-Tetrachloroethane	-1.0000	0.2998	0.0	0	0.00
70	C	Ethylbenzene	-1.0000	0.3040	0.0	0	0.00
71	T	m-,p-Xylene	-1.0000	0.6051	0.0	0	0.01
72	T	o-Xylene	-1.0000	0.2994	0.0	0	0.00
73	T	Styrene	-1.0000	0.2961	0.0	0	0.00
74	P	Bromoform	-1.0000	0.0798	0.0	0	-0.01
75	T	Isopropylbenzene	-1.0000	0.2931	0.0	0	0.00
76	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
77	P	1,1,2,2-Tetrachloroethane	-1.0000	0.2922	0.0	0	0.00
78	S	p-Bromofluorobenzene	-1.0000	0.0730	0.0	0	0.00
79	T	1,2,3-Trichloropropane	-1.0000	0.0000	0.0	0	-15.70#
80	T	trans-1,4-Dichloro-2-Butene	-1.0000	0.0000	0.0	0	-15.74#
81	T	n-Propylbenzene	-1.0000	0.3233	0.0	0	0.00
82	T	Bromobenzene	0.3000	0.2618	12.7	100	0.00
83	T	1,3,5-Trimethylbenzene	-1.0000	0.3032	0.0	0	-0.01
84	T	2-Chlorotoluene	-1.0000	0.3300	0.0	0	0.00
85	T	4-Chlorotoluene	-1.0000	0.3295	0.0	0	0.00
86	T	a-Methylstyrene	-1.0000	0.2430	0.0	0	0.00
87	T	tert-Butylbenzene	-1.0000	0.2541	0.0	0	-0.01
88	T	1,2,4-Trimethylbenzene	-1.0000	0.3088	0.0	0	0.00
89	T	sec-Butylbenzene	-1.0000	0.3122	0.0	0	0.00
90	T	p-Isopropyltoluene	-1.0000	0.2893	0.0	0	0.00
91	T	1,3-Dichlorobenzene	-1.0000	0.3051	0.0	0	0.00
92	T	1,4-Dichlorobenzene	0.3000	0.2990	0.3	100	0.00
93	T	n-Butylbenzene	-1.0000	0.3274	0.0	0	0.00
94	T	1,2-Dichlorobenzene	0.3000	0.2985	0.5	100	-0.01
95	T	1,2-Dibromo-3-Chloropropane	-1.0000	0.0000	0.0	0	-18.49#
96	T	1,2,4-Trichlorobenzene	-1.0000	0.3011	0.0	0	0.00
97	T	Hexachlorobutadiene	-1.0000	0.2675	0.0	0	0.00
98	T	Naphthalene	-1.0000	0.2561	0.0	0	0.00
99	T	1,2,3-Trichlorobenzene	0.3000	0.2854	4.9	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
11M11837.D 8260WT.M Sat May 14 18:41:07 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11838.D Vial: 4
Acq On : 13 May 2016 15:12 Operator: JDS
Sample : WG568769-03 0.4ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04:23 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	515473	25.00 ug/L	-0.01
56) Chlorobenzene-d5	14.25	117	435266	25.00 ug/L	-0.01
76) 1,4-Dichlorobenzene-d4	17.06	152	244818	25.00 ug/L	-0.01

System Monitoring Compounds					
37) Dibromofluoromethane	0.00	111	0	0.0000 ug/L	
Spiked Amount 25.000	Range 86 - 118		Recovery =	0.00%#	
43) 1,2-Dichloroethane-d4	0.00	65	0	0.0000 ug/L	
Spiked Amount 25.000	Range 80 - 120		Recovery =	0.00%#	
57) Toluene-d8	0.00	98	0	0.0000 ug/L	
Spiked Amount 25.000	Range 88 - 110		Recovery =	0.00%#	
78) p-Bromofluorobenzene	15.64	95	236	0.0295 ug/L	-0.01
Spiked Amount 25.000	Range 86 - 115		Recovery =	0.12%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	2765	0.3656 ug/L #	66
3) Chloromethane	3.72	50	4021	0.5854 ug/L #	77
4) Vinyl Chloride	3.96	62	2267	0.4134 ug/L #	64
5) 1,3-Butadiene	4.01	54	2044	0.3956 ug/L #	18
6) Bromomethane	4.85	94	1571	0.4511 ug/L	89
7) Chloroethane	5.01	64	1496	0.4029 ug/L #	45
8) Trichlorofluoromethane	5.49	101	4241	0.4175 ug/L	96
10) Isoprene	6.03	67	2394	0.3408 ug/L	90
12) 1,1,2-Trichloro-1,2,2-Trif	6.28	101	1439	0.2811 ug/L #	23
13) Acetone	6.33	43	1125	0.7299 ug/L #	45
14) 1,1-Dichloroethene	6.56	61	3786	0.3990 ug/L	77
16) Dimethyl Sulfide	6.80	62	1622	0.3769 ug/L	59
18) Methyl acetate	7.01	43	21364	4.0629 ug/L #	70
19) Methylene Chloride	7.32	84	2266	0.4265 ug/L	97
20) Carbon Disulfide	7.35	76	6628	0.3991 ug/L #	82
22) Methyl Tert Butyl Ether	7.52	73	4808	0.3486 ug/L #	74
23) trans-1,2-Dichloroethene	7.74	96	2372	0.4368 ug/L	93
24) n-Hexane	7.83	57	4212	0.4732 ug/L #	73
26) Vinyl Acetate	8.34	43	434	0.1315 ug/L #	78
27) 1,1-Dichloroethane	8.34	63	3822	0.3576 ug/L #	89
29) 2-Butanone	8.88	43	484	0.2099 ug/L #	60
31) 2,2-Dichloropropane	9.09	77	2680	0.3472 ug/L #	55
32) cis-1,2-Dichloroethene	9.15	96	2277	0.3777 ug/L	91
33) Chloroform	9.35	83	3981	0.3988 ug/L	89
35) Bromochloromethane	9.57	130	952	0.2468 ug/L	57
36) Tetrahydrofuran	9.59	42	212	0.1340 ug/L #	46
38) 1,1,1-Trichloroethane	9.85	97	3917	0.4058 ug/L #	84
39) Cyclohexane	9.88	56	3935	0.3371 ug/L #	80
40) 1,1-Dichloropropene	10.04	75	2817	0.3862 ug/L	82
41) Carbon Tetrachloride	10.18	117	3373	0.3702 ug/L #	88
44) 1,2-Dichloroethane	10.35	62	3201	0.3737 ug/L #	78
45) Benzene	10.39	78	8729	0.4184 ug/L	90
46) Trichloroethene	11.08	130	2537	0.3731 ug/L	93
47) Methylcyclohexane	11.17	83	2742	0.3325 ug/L	90
48) 1,2-Dichloropropane	11.29	63	2133	0.3600 ug/L	84
50) Bromodichloromethane	11.58	83	2857	0.3638 ug/L #	84
51) Dibromomethane	11.65	93	972	0.3128 ug/L	81
52) 2-Chloroethyl Vinyl Ether	11.85	63	791	0.2560 ug/L #	50
54) cis-1,3-Dichloropropene	12.17	75	2689	0.3253 ug/L	99
55) Dimethyl Disulfide	12.42	79	1113	0.2158 ug/L	74

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11838.D Vial: 4
Acq On : 13 May 2016 15:12 Operator: JDS
Sample : WG568769-03 0.4ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04:23 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
58)	Toluene	12.56	91	9113	0.4043 ug/L		99
59)	Ethyl Methacrylate	12.65	69	1556	0.2916 ug/L		96
60)	trans-1,3-Dichloropropene	12.73	75	2303	0.3119 ug/L		86
61)	1,1,2-Trichloroethane	12.94	97	1549	0.3672 ug/L		85
62)	2-Hexanone	12.86	43	840	0.2415 ug/L	#	11
63)	1,3-Dichloropropane	13.21	76	2345	0.3346 ug/L		82
64)	Tetrachloroethene	13.34	164	1905	0.3745 ug/L		88
65)	Dibromochloromethane	13.59	129	1972	0.3238 ug/L		89
66)	1,2-Dibromoethane	13.82	107	1617	0.3776 ug/L		69
67)	1-Chlorohexane	13.89	91	2578	0.3487 ug/L		77
68)	Chlorobenzene	14.29	112	6458	0.3896 ug/L		93
69)	1,1,1,2-Tetrachloroethane	14.31	131	2374	0.3736 ug/L		98
70)	Ethylbenzene	14.31	106	3158	0.3754 ug/L		79
71)	m-,p-Xylene	14.39	106	8621	0.8613 ug/L		93
72)	o-Xylene	14.93	106	3752	0.3758 ug/L		96
73)	Styrene	14.95	104	6027	0.3558 ug/L		90
74)	Bromoform	15.44	173	931	0.2454 ug/L		87
75)	Isopropylbenzene	15.31	105	10551	0.4179 ug/L		95
77)	1,1,2,2-Tetrachloroethane	15.52	83	1129	0.2796 ug/L		84
79)	1,2,3-Trichloropropane	15.71	110	200	0.1341 ug/L	#	28
81)	n-Propylbenzene	15.79	91	11244	0.4087 ug/L		98
82)	Bromobenzene	15.91	156	3048	0.3991 ug/L		95
83)	1,3,5-Trimethylbenzene	15.96	105	9158	0.4381 ug/L		92
84)	2-Chlorotoluene	16.05	91	8455	0.4164 ug/L		96
85)	4-Chlorotoluene	16.09	91	6979	0.4189 ug/L		92
86)	a-Methylstyrene	16.34	118	4052	0.3488 ug/L		74
87)	tert-Butylbenzene	16.40	134	1479	0.3224 ug/L		63
88)	1,2,4-Trimethylbenzene	16.45	105	8457	0.3953 ug/L		95
89)	sec-Butylbenzene	16.65	105	10479	0.4236 ug/L		93
90)	p-Isopropyltoluene	16.79	119	9613	0.4269 ug/L		93
91)	1,3-Dichlorobenzene	16.98	146	5826	0.4107 ug/L		93
92)	1,4-Dichlorobenzene	17.10	146	6273	0.4331 ug/L	#	74
93)	n-Butylbenzene	17.28	91	8021	0.3956 ug/L	#	91
94)	1,2-Dichlorobenzene	17.57	146	4937	0.3691 ug/L		97
96)	1,2,4-Trichlorobenzene	19.55	180	3632	0.3655 ug/L		95
97)	Hexachlorobutadiene	19.69	225	1539	0.3812 ug/L		96
98)	Naphthalene	19.90	128	7383	0.3882 ug/L		97
99)	1,2,3-Trichlorobenzene	20.19	180	3748	0.4028 ug/L		84

(#) = qualifier out of range (m) = manual integration
11M11838.D 8260WT.M Fri May 13 17:04:24 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11838.D Vial: 4
Acq On : 13 May 2016 15:12 Operator: JDS
Sample : WG568769-03 0.4ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11839.D Vial: 5
Acq On : 13 May 2016 15:43 Operator: JDS
Sample : WG568769-04 1.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04:25 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	523425	25.00 ug/L	-0.01
56) Chlorobenzene-d5	14.25	117	440574	25.00 ug/L	-0.01
76) 1,4-Dichlorobenzene-d4	17.06	152	249989	25.00 ug/L	-0.01

System Monitoring Compounds					
37) Dibromofluoromethane	9.63	111	1729	0.2910 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery =	1.16%#	
43) 1,2-Dichloroethane-d4	10.23	65	3301	0.4585 ug/L	-0.01
Spiked Amount 25.000	Range 80 - 120		Recovery =	1.84%#	
57) Toluene-d8	12.47	98	9307	0.4494 ug/L	-0.01
Spiked Amount 25.000	Range 88 - 110		Recovery =	1.80%#	
78) p-Bromofluorobenzene	15.64	95	4047	0.4956 ug/L	-0.01
Spiked Amount 25.000	Range 86 - 115		Recovery =	2.00%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	7215	0.9394 ug/L	98
3) Chloromethane	3.72	50	7292	1.0455 ug/L	80
4) Vinyl Chloride	3.96	62	5040	0.9051 ug/L	92
5) 1,3-Butadiene	4.02	54	5876	1.1198 ug/L #	76
6) Bromomethane	4.85	94	3488	0.9862 ug/L	94
7) Chloroethane	4.99	64	3425	0.9083 ug/L #	57
8) Trichlorofluoromethane	5.49	101	10960	1.0625 ug/L	95
9) Diethyl ether	6.01	59	19830	4.2253 ug/L	90
10) Isoprene	6.04	67	6121	0.8582 ug/L	94
11) Acrolein	6.25	56	697	1.2730 ug/L	65
12) 1,1,2-Trichloro-1,2,2-Trif	6.27	101	5456	1.0497 ug/L	90
13) Acetone	6.34	43	1782	1.1386 ug/L #	45
14) 1,1-Dichloroethene	6.55	61	9456	0.9814 ug/L	91
15) Tert-Butyl Alcohol	6.67	59	2848	6.8484 ug/L #	67
16) Dimethyl Sulfide	6.80	62	3631	0.8310 ug/L	93
18) Methyl acetate	7.01	43	20953	3.9242 ug/L #	70
19) Methylene Chloride	7.31	84	5001	0.9270 ug/L	97
20) Carbon Disulfide	7.36	76	15438	0.9154 ug/L	99
21) Acrylonitrile	7.50	53	3780	1.8795 ug/L	89
22) Methyl Tert Butyl Ether	7.53	73	11891	0.8490 ug/L	90
23) trans-1,2-Dichloroethene	7.75	96	4815	0.8732 ug/L	95
24) n-Hexane	7.82	57	9213	1.0194 ug/L #	81
25) Diisopropyl ether	8.15	45	108233	4.5267 ug/L	99
26) Vinyl Acetate	8.31	43	1723	0.5141 ug/L #	78
27) 1,1-Dichloroethane	8.34	63	10818	0.9968 ug/L	95
28) Ethyl-Tert-Butyl ether	8.70	59	85288	4.3548 ug/L	99
29) 2-Butanone	8.88	43	1368	0.5844 ug/L #	60
30) Propionitrile	8.99	54	2406	3.5185 ug/L #	60
31) 2,2-Dichloropropane	9.09	77	7787	0.9935 ug/L	90
32) cis-1,2-Dichloroethene	9.15	96	5567	0.9095 ug/L	94
33) Chloroform	9.35	83	10278	1.0139 ug/L	96
34) 1-Bromopropane	9.48	122	295	0.2934 ug/L	49
35) Bromochloromethane	9.57	130	2941	0.7509 ug/L	88
36) Tetrahydrofuran	9.60	42	7790	4.8493 ug/L	92
38) 1,1,1-Trichloroethane	9.85	97	10000	1.0202 ug/L #	90
39) Cyclohexane	9.88	56	10164	0.8574 ug/L	98
40) 1,1-Dichloropropene	10.04	75	7321	0.9884 ug/L	92
41) Carbon Tetrachloride	10.18	117	9558	1.0331 ug/L	98
42) Tert-Amyl-Methyl ether	10.13	73	59107	4.2417 ug/L	97
44) 1,2-Dichloroethane	10.34	62	8181	0.9406 ug/L	99

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11839.D Vial: 5
Acq On : 13 May 2016 15:43 Operator: JDS
Sample : WG568769-04 1.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04:25 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
45)	Benzene	10.38	78	19924	0.9404 ug/L		95
46)	Trichloroethene	11.08	130	6636	0.9611 ug/L		92
47)	Methylcyclohexane	11.18	83	7740	0.9243 ug/L		94
48)	1,2-Dichloropropane	11.29	63	5477	0.9103 ug/L		99
49)	1,4-Dioxane	11.55	88	187	5.1264 ug/L	#	10
50)	Bromodichloromethane	11.57	83	7238	0.9076 ug/L		94
51)	Dibromomethane	11.65	93	2839	0.8997 ug/L		97
52)	2-Chloroethyl Vinyl Ether	11.84	63	1934	0.6164 ug/L	#	72
53)	4-Methyl-2-Pentanone	11.87	58	1016	0.5282 ug/L	#	42
54)	cis-1,3-Dichloropropene	12.17	75	7078	0.8432 ug/L		99
55)	Dimethyl Disulfide	12.42	79	3514	0.6710 ug/L		93
58)	Toluene	12.56	91	22074	0.9676 ug/L		96
59)	Ethyl Methacrylate	12.64	69	4130	0.7646 ug/L		88
60)	trans-1,3-Dichloropropene	12.73	75	5897	0.7891 ug/L		90
61)	1,1,2-Trichloroethane	12.94	97	3547	0.8307 ug/L		96
62)	2-Hexanone	12.87	43	2461	0.6990 ug/L	#	63
63)	1,3-Dichloropropane	13.21	76	5965	0.8410 ug/L		87
64)	Tetrachloroethene	13.34	164	5050	0.9809 ug/L		94
65)	Dibromochloromethane	13.59	129	5237	0.8497 ug/L		90
66)	1,2-Dibromoethane	13.82	107	3214	0.7414 ug/L		91
67)	1-Chlorohexane	13.89	91	7025	0.9387 ug/L		91
68)	Chlorobenzene	14.29	112	16160	0.9630 ug/L		95
69)	1,1,1,2-Tetrachloroethane	14.32	131	6361	0.9889 ug/L		97
70)	Ethylbenzene	14.31	106	8050	0.9455 ug/L		93
71)	m-,p-Xylene	14.39	106	20545	2.0278 ug/L		95
72)	o-Xylene	14.92	106	9683	0.9581 ug/L		96
73)	Styrene	14.96	104	14946	0.8716 ug/L		98
74)	Bromoform	15.43	173	3024	0.7876 ug/L		97
75)	Isopropylbenzene	15.31	105	25656	1.0039 ug/L		92
77)	1,1,2,2-Tetrachloroethane	15.52	83	3569	0.8655 ug/L		89
79)	1,2,3-Trichloropropane	15.71	110	872	0.5724 ug/L		84
80)	trans-1,4-Dichloro-2-Buten	15.74	53	1048	0.6163 ug/L	#	44
81)	n-Propylbenzene	15.79	91	27998	0.9967 ug/L		98
82)	Bromobenzene	15.92	156	8101	1.0388 ug/L		89
83)	1,3,5-Trimethylbenzene	15.96	105	20563	0.9633 ug/L		98
84)	2-Chlorotoluene	16.05	91	20962	1.0111 ug/L		96
85)	4-Chlorotoluene	16.09	91	17553	1.0318 ug/L		100
86)	a-Methylstyrene	16.34	118	10002	0.8431 ug/L		97
87)	tert-Butylbenzene	16.40	134	4396	0.9384 ug/L		84
88)	1,2,4-Trimethylbenzene	16.44	105	21945	1.0045 ug/L		96
89)	sec-Butylbenzene	16.65	105	25927	1.0263 ug/L		100
90)	p-Isopropyltoluene	16.79	119	22467	0.9771 ug/L		96
91)	1,3-Dichlorobenzene	16.98	146	14883	1.0274 ug/L		96
92)	1,4-Dichlorobenzene	17.10	146	15106	1.0213 ug/L	#	67
93)	n-Butylbenzene	17.28	91	20991	1.0139 ug/L		98
94)	1,2-Dichlorobenzene	17.57	146	13331	0.9762 ug/L		99
95)	1,2-Dibromo-3-Chloropropan	18.49	75	493	0.5750 ug/L		75
96)	1,2,4-Trichlorobenzene	19.55	180	9030	0.8899 ug/L		96
97)	Hexachlorobutadiene	19.69	225	3810	0.9242 ug/L		95
98)	Naphthalene	19.90	128	16333	0.8410 ug/L		99
99)	1,2,3-Trichlorobenzene	20.19	180	8645	0.9098 ug/L		88

(#) = qualifier out of range (m) = manual integration
11M11839.D 8260WT.M Fri May 13 17:04:25 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11839.D Vial: 5
Acq On : 13 May 2016 15:43 Operator: JDS
Sample : WG568769-04 1.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\051316\11M11839.D Vial: 5
Acq On : 13 May 2016 15:43 Operator: JDS
Sample : WG568769-04 1.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:40:10 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	1.0000	0.8798	12.0	100	0.00
3	P	Chloromethane	1.0000	1.1102	-11.0	100	0.00
4	C	Vinyl Chloride	1.0000	0.9240	7.6	100	0.00
5	T	1,3-Butadiene	-1.0000	1.0834	0.0	100	0.02
6	T	Bromomethane	1.0000	0.9301	7.0	100	0.00
7	T	Chloroethane	1.0000	0.9773	2.3	100	-0.01
8	T	Trichlorofluoromethane	1.0000	1.0081	-0.8	100	0.01
9	T	Diethyl ether	5.0000	4.9318	1.4	100	0.00
10	T	Isoprene	-1.0000	0.9134	0.0	100	0.00
11	T	Acrolein	2.5000	2.7854	-11.4	100	0.01
12	T	1,1,2-Trichloro-1,2,2-Trifl	1.0000	1.0342	-3.4	100	0.02
13	T	Acetone	-1.0000	1.4337	0.0	0	0.00
14	C	1,1-Dichloroethene	1.0000	0.9868	1.3	100	-0.01
15	T	Tert-Butyl Alcohol	10.0000	8.2841	17.2	100	0.01
16	T	Dimethyl Sulfide	-1.0000	0.9488	0.0	100	-0.01
17	T	Iodomethane	-1.0000	2.2962	0.0	0	0.01
18	T	Methyl acetate	-1.0000	4.2835	0.0	0	-0.05
19	T	Methylene Chloride	1.0000	1.0084	-0.8	100	0.00
20	T	Carbon Disulfide	1.0000	0.9624	3.8	100	0.00
21	T	Acrylonitrile	2.5000	2.2683	9.3	100	0.01
22	T	Methyl Tert Butyl Ether	1.0000	0.9614	3.9	100	0.01
23	T	trans-1,2-Dichloroethene	1.0000	0.9079	9.2	100	0.01
24	T	n-Hexane	-1.0000	1.0430	0.0	100	0.00
25	T	Diisopropyl ether	5.0000	4.9106	1.8	100	0.00
26	T	Vinyl Acetate	-1.0000	2.6242	0.0	0	0.00
27	P	1,1-Dichloroethane	1.0000	1.0166	-1.7	100	0.00
28	T	Ethyl-Tert-Butyl ether	5.0000	4.7543	4.9	100	0.00
29	T	2-Butanone	-1.0000	0.7280	0.0	0	0.01
30	T	Propionitrile	5.0000	4.3718	12.6	100	0.01
31	T	2,2-Dichloropropane	1.0000	0.9353	6.5	100	0.00
32	T	cis-1,2-Dichloroethene	1.0000	0.9520	4.8	100	0.00
33	C	Chloroform	1.0000	0.9836	1.6	100	0.00
34	T	1-Bromopropane	1.0000	0.9106	8.9	100	0.00
35	T	Bromochloromethane	1.0000	0.8469	15.3	100	0.00
36	T	Tetrahydrofuran	5.0000	5.8710	-17.4	100	0.00
37	S	Dibromofluoromethane	-1.0000	0.3018	0.0	100	0.00
38	T	1,1,1-Trichloroethane	1.0000	0.9725	2.8	100	0.00
39	T	Cyclohexane	1.0000	0.9069	9.3	100	0.00
40	T	1,1-Dichloropropene	1.0000	0.9971	0.3	100	0.00
41	T	Carbon Tetrachloride	1.0000	0.9518	4.8	100	0.00
42	T	Tert-Amyl-Methyl ether	5.0000	4.7417	5.2	100	0.00
43	S	1,2-Dichloroethane-d4	0.5000	0.5054	-1.1	100	0.00
44	T	1,2-Dichloroethane	1.0000	0.9639	3.6	100	0.00
45	T	Benzene	1.0000	0.9912	0.9	100	0.00
46	T	Trichloroethene	1.0000	1.0157	-1.6	100	-0.01
47	T	Methylcyclohexane	1.0000	0.9600	4.0	100	0.01
48	C	1,2-Dichloropropane	1.0000	0.9844	1.6	100	0.00
49	T	1,4-Dioxane	-1.0000	28.5532	0.0	0	0.00
50	T	Bromodichloromethane	1.0000	0.9266	7.3	100	0.00
51	T	Dibromomethane	1.0000	0.9792	2.1	100	0.00
52	T	2-Chloroethyl Vinyl Ether	-1.0000	0.7053	0.0	100	0.00
53	T	4-Methyl-2-Pentanone	-1.0000	0.6569	0.0	0	0.00
54	T	cis-1,3-Dichloropropene	1.0000	0.9060	9.4	100	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\051316\11M11839.D Vial: 5
Acq On : 13 May 2016 15:43 Operator: JDS
Sample : WG568769-04 1.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:40:10 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	Dimethyl Disulfide	-1.0000	0.7475	0.0	100	0.00
56	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
57	S	Toluene-d8	-1.0000	0.4862	0.0	100	0.00
58	C	Toluene	1.0000	0.9897	1.0	100	0.00
59	T	Ethyl Methacrylate	1.0000	0.8672	13.3	100	-0.01
60	T	trans-1,3-Dichloropropene	1.0000	0.8237	17.6	100	0.00
61	T	1,1,2-Trichloroethane	1.0000	0.8969	10.3	100	0.00
62	T	2-Hexanone	-1.0000	0.8473	0.0	100	0.01
63	T	1,3-Dichloropropane	1.0000	0.9392	6.1	100	0.00
64	T	Tetrachloroethene	1.0000	0.9737	2.6	100	0.00
65	T	Dibromochloromethane	1.0000	0.8681	13.2	100	0.00
66	T	1,2-Dibromoethane	1.0000	0.8111	18.9	100	0.00
67	T	1-Chlorohexane	1.0000	0.9554	4.5	100	0.00
68	P	Chlorobenzene	1.0000	0.9890	1.1	100	0.00
69	T	1,1,1,2-Tetrachloroethane	1.0000	0.9621	3.8	100	0.00
70	C	Ethylbenzene	1.0000	0.9608	3.9	100	0.00
71	T	m-,p-Xylene	2.0000	2.0170	-0.9	100	0.00
72	T	o-Xylene	1.0000	0.9768	2.3	100	0.00
73	T	Styrene	1.0000	0.9126	8.7	100	0.01
74	P	Bromoform	1.0000	0.8271	17.3	100	0.00
75	T	Isopropylbenzene	1.0000	0.9787	2.1	100	0.00
76	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
77	P	1,1,2,2-Tetrachloroethane	1.0000	0.9168	8.3	100	0.00
78	S	p-Bromofluorobenzene	-1.0000	0.5252	0.0	100	0.00
79	T	1,2,3-Trichloropropane	1.0000	0.9573	4.3	100	0.01
80	T	trans-1,4-Dichloro-2-Butene	1.0000	1.2934	-29.3#	100	0.00
81	T	n-Propylbenzene	1.0000	0.9671	3.3	100	0.00
82	T	Bromobenzene	1.0000	1.0663	-6.6	100	0.00
83	T	1,3,5-Trimethylbenzene	1.0000	0.9307	6.9	100	0.00
84	T	2-Chlorotoluene	1.0000	1.0337	-3.4	100	0.00
85	T	4-Chlorotoluene	1.0000	0.9664	3.4	100	0.00
86	T	a-Methylstyrene	1.0000	0.8582	14.2	100	0.00
87	T	tert-Butylbenzene	1.0000	0.9392	6.1	100	0.00
88	T	1,2,4-Trimethylbenzene	1.0000	0.9755	2.5	100	0.00
89	T	sec-Butylbenzene	1.0000	0.9774	2.3	100	0.00
90	T	p-Isopropyltoluene	1.0000	0.9221	7.8	100	0.00
91	T	1,3-Dichlorobenzene	1.0000	1.0144	-1.4	100	0.00
92	T	1,4-Dichlorobenzene	1.0000	1.0157	-1.6	100	0.00
93	T	n-Butylbenzene	1.0000	0.9776	2.2	100	0.00
94	T	1,2-Dichlorobenzene	1.0000	0.9974	0.3	100	0.00
95	T	1,2-Dibromo-3-Chloropropane	-1.0000	1.1961	0.0	100	0.00
96	T	1,2,4-Trichlorobenzene	1.0000	0.9233	7.7	100	0.00
97	T	Hexachlorobutadiene	1.0000	0.8843	11.6	100	0.00
98	T	Naphthalene	1.0000	0.8997	10.0	100	0.00
99	T	1,2,3-Trichlorobenzene	1.0000	0.9631	3.7	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
11M11839.D 8260WT.M Sat May 14 18:41:49 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11840.D Vial: 6
Acq On : 13 May 2016 16:15 Operator: JDS
Sample : WG568769-05 2.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04:26 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	504626	25.00 ug/L	-0.01
56) Chlorobenzene-d5	14.25	117	418265	25.00 ug/L	-0.01
76) 1,4-Dichlorobenzene-d4	17.06	152	241104	25.00 ug/L	-0.01

System Monitoring Compounds					
37) Dibromofluoromethane	9.63	111	5067	0.8846 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery =	3.52%#	
43) 1,2-Dichloroethane-d4	10.23	65	6143	0.8851 ug/L	-0.01
Spiked Amount 25.000	Range 80 - 120		Recovery =	3.56%#	
57) Toluene-d8	12.47	98	18425	0.9371 ug/L	-0.01
Spiked Amount 25.000	Range 88 - 110		Recovery =	3.76%#	
78) p-Bromofluorobenzene	15.64	95	7442	0.9449 ug/L	-0.01
Spiked Amount 25.000	Range 86 - 115		Recovery =	3.76%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	14866	2.0077 ug/L	98
3) Chloromethane	3.72	50	14641	2.1775 ug/L	94
4) Vinyl Chloride	3.96	62	10045	1.8711 ug/L	99
5) 1,3-Butadiene	4.01	54	10590	2.0934 ug/L	82
6) Bromomethane	4.85	94	7024	2.0600 ug/L	94
7) Chloroethane	4.99	64	6661	1.8324 ug/L	97
8) Trichlorofluoromethane	5.49	101	21690	2.1811 ug/L	99
9) Diethyl ether	6.01	59	96350	21.2945 ug/L	97
10) Isoprene	6.04	67	12682	1.8443 ug/L	94
11) Acrolein	6.23	56	5531	10.4784 ug/L	91
12) 1,1,2-Trichloro-1,2,2-Trif	6.23	101	9909	1.9774 ug/L	96
13) Acetone	6.35	43	3346	2.2176 ug/L #	68
14) 1,1-Dichloroethene	6.56	61	18584	2.0007 ug/L	98
15) Tert-Butyl Alcohol	6.66	59	15203	37.9194 ug/L #	93
16) Dimethyl Sulfide	6.81	62	7308	1.7348 ug/L	98
17) Iodomethane	7.06	142	1747	0.4335 ug/L #	49
18) Methyl acetate	7.01	43	17794	3.4567 ug/L #	75
19) Methylene Chloride	7.31	84	10330	1.9861 ug/L	99
20) Carbon Disulfide	7.36	76	30745	1.8910 ug/L	99
21) Acrylonitrile	7.49	53	17499	9.0252 ug/L	91
22) Methyl Tert Butyl Ether	7.52	73	22031	1.6315 ug/L	97
23) trans-1,2-Dichloroethene	7.74	96	10660	2.0051 ug/L	95
24) n-Hexane	7.82	57	17125	1.9654 ug/L #	94
25) Diisopropyl ether	8.15	45	542272	23.5246 ug/L	99
26) Vinyl Acetate	8.31	43	6813	2.1084 ug/L #	84
27) 1,1-Dichloroethane	8.34	63	21278	2.0336 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	431080	22.8311 ug/L	100
29) 2-Butanone	8.87	43	3175	1.4068 ug/L #	60
30) Propionitrile	8.99	54	12529	19.0049 ug/L	98
31) 2,2-Dichloropropane	9.09	77	17576	2.3259 ug/L	94
32) cis-1,2-Dichloroethene	9.15	96	11652	1.9746 ug/L	96
33) Chloroform	9.35	83	20769	2.1252 ug/L	96
34) 1-Bromopropane	9.48	122	1648	1.7003 ug/L	70
35) Bromochloromethane	9.57	130	6812	1.8040 ug/L	100
36) Tetrahydrofuran	9.60	42	30834	19.9093 ug/L	96
38) 1,1,1-Trichloroethane	9.84	97	20518	2.1713 ug/L #	94
39) Cyclohexane	9.88	56	22087	1.9325 ug/L	98
40) 1,1-Dichloropropene	10.04	75	13954	1.9542 ug/L	98
41) Carbon Tetrachloride	10.18	117	21054	2.3604 ug/L	98
42) Tert-Amyl-Methyl ether	10.13	73	300593	22.3753 ug/L	99

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11840.D Vial: 6
Acq On : 13 May 2016 16:15 Operator: JDS
Sample : WG568769-05 2.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04:26 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Compound	R.T.	QIon	Response	Conc Unit	Qvalue
44) 1,2-Dichloroethane	10.34	62	16928	2.0188 ug/L	98
45) Benzene	10.38	78	41004	2.0075 ug/L	97
46) Trichloroethene	11.08	130	13283	1.9954 ug/L	96
47) Methylcyclohexane	11.17	83	15711	1.9461 ug/L	95
48) 1,2-Dichloropropane	11.29	63	11010	1.8980 ug/L	91
49) 1,4-Dioxane	11.55	88	898	25.5347 ug/L	88
50) Bromodichloromethane	11.57	83	15058	1.9584 ug/L	100
51) Dibromomethane	11.66	93	5461	1.7952 ug/L	97
52) 2-Chloroethyl Vinyl Ether	11.84	63	4894	1.6180 ug/L	83
53) 4-Methyl-2-Pentanone	11.87	58	2173	1.1717 ug/L #	65
54) cis-1,3-Dichloropropene	12.17	75	14440	1.7842 ug/L	91
55) Dimethyl Disulfide	12.42	79	7378	1.4613 ug/L	97
58) Toluene	12.56	91	43051	1.9877 ug/L	98
59) Ethyl Methacrylate	12.64	69	7792	1.5195 ug/L	95
60) trans-1,3-Dichloropropene	12.73	75	12991	1.8310 ug/L	98
61) 1,1,2-Trichloroethane	12.93	97	8138	2.0075 ug/L	91
62) 2-Hexanone	12.86	43	4715	1.4106 ug/L #	61
63) 1,3-Dichloropropane	13.21	76	11740	1.7434 ug/L	95
64) Tetrachloroethene	13.34	164	10610	2.1708 ug/L	90
65) Dibromochloromethane	13.59	129	11236	1.9202 ug/L	96
66) 1,2-Dibromoethane	13.82	107	7613	1.8499 ug/L	99
67) 1-Chlorohexane	13.89	91	13853	1.9499 ug/L	96
68) Chlorobenzene	14.29	112	33381	2.0954 ug/L	99
69) 1,1,1,2-Tetrachloroethane	14.32	131	13541	2.2175 ug/L	97
70) Ethylbenzene	14.31	106	17116	2.1175 ug/L	98
71) m-,p-Xylene	14.39	106	40469	4.2073 ug/L	95
72) o-Xylene	14.92	106	19196	2.0006 ug/L	99
73) Styrene	14.96	104	30019	1.8440 ug/L	92
74) Bromoform	15.43	173	6309	1.7308 ug/L	96
75) Isopropylbenzene	15.31	105	52419	2.1605 ug/L	98
77) 1,1,2,2-Tetrachloroethane	15.51	83	7118	1.7898 ug/L	100
79) 1,2,3-Trichloropropane	15.70	110	2487	1.6927 ug/L	98
80) trans-1,4-Dichloro-2-Buten	15.74	53	2265	1.3810 ug/L	68
81) n-Propylbenzene	15.79	91	59353	2.1908 ug/L	100
82) Bromobenzene	15.91	156	15393	2.0465 ug/L	97
83) 1,3,5-Trimethylbenzene	15.95	105	43641	2.1197 ug/L	98
84) 2-Chlorotoluene	16.05	91	40923	2.0467 ug/L	96
85) 4-Chlorotoluene	16.09	91	37360	2.2771 ug/L	97
86) a-Methylstyrene	16.34	118	20567	1.7976 ug/L	92
87) tert-Butylbenzene	16.40	134	9253	2.0481 ug/L	89
88) 1,2,4-Trimethylbenzene	16.44	105	43857	2.0815 ug/L	99
89) sec-Butylbenzene	16.65	105	52733	2.1643 ug/L	99
90) p-Isopropyltoluene	16.79	119	49990	2.2541 ug/L	98
91) 1,3-Dichlorobenzene	16.98	146	29674	2.1240 ug/L	99
92) 1,4-Dichlorobenzene	17.10	146	30258	2.1211 ug/L	87
93) n-Butylbenzene	17.28	91	44260	2.2167 ug/L	98
94) 1,2-Dichlorobenzene	17.57	146	27619	2.0969 ug/L	95
95) 1,2-Dibromo-3-Chloropropan	18.49	75	1094	1.3229 ug/L	60
96) 1,2,4-Trichlorobenzene	19.55	180	18455	1.8857 ug/L	98
97) Hexachlorobutadiene	19.69	225	8691	2.1858 ug/L	89
98) Naphthalene	19.90	128	34041	1.8174 ug/L #	95
99) 1,2,3-Trichlorobenzene	20.19	180	17703	1.9318 ug/L	99

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11840.D Vial: 6
Acq On : 13 May 2016 16:15 Operator: JDS
Sample : WG568769-05 2.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:04 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 11:37:47 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\051316\11M11840.D Vial: 6
Acq On : 13 May 2016 16:15 Operator: JDS
Sample : WG568769-05 2.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	2.0000	1.8802	6.0	100	0.00
3	P	Chloromethane	2.0000	2.3121	-15.6	100	0.00
4	C	Vinyl Chloride	2.0000	1.9102	4.5	100	0.00
5	T	1,3-Butadiene	-1.0000	2.0253	0.0	100	0.01
6	T	Bromomethane	2.0000	1.9427	2.9	100	0.00
7	T	Chloroethane	2.0000	1.9714	1.4	100	-0.01
8	T	Trichlorofluoromethane	2.0000	2.0694	-3.5	100	0.01
9	T	Diethyl ether	25.0000	24.8555	0.6	100	0.00
10	T	Isoprene	-1.0000	1.9630	0.0	100	0.00
11	T	Acrolein	12.5000	12.0322	3.7	100	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	2.0000	1.9483	2.6	100	-0.01
13	T	Acetone	-1.0000	2.7924	0.0	100	0.01
14	C	1,1-Dichloroethene	2.0000	2.0117	-0.6	100	0.00
15	T	Tert-Butyl Alcohol	50.0000	45.8688	8.3	100	0.00
16	T	Dimethyl Sulfide	-1.0000	1.9809	0.0	100	0.00
17	T	Iodomethane	2.0000	2.6580	-32.9#	100	0.00
18	T	Methyl acetate	2.0000	0.8555	57.2#	100	0.00
19	T	Methylene Chloride	2.0000	2.1606	-8.0	100	0.00
20	T	Carbon Disulfide	2.0000	1.9879	0.6	100	0.00
21	T	Acrylonitrile	12.5000	10.8919	12.9	100	0.00
22	T	Methyl Tert Butyl Ether	2.0000	1.8475	7.6	100	0.00
23	T	trans-1,2-Dichloroethene	2.0000	2.0850	-4.2	100	0.00
24	T	n-Hexane	2.0000	2.0110	-0.6	100	0.00
25	T	Diisopropyl ether	25.0000	25.5200	-2.1	100	0.00
26	T	Vinyl Acetate	-1.0000	3.2184	0.0	100	0.00
27	P	1,1-Dichloroethane	2.0000	2.0741	-3.7	100	0.00
28	T	Ethyl-Tert-Butyl ether	25.0000	24.9254	0.3	100	0.00
29	T	2-Butanone	-1.0000	1.7527	0.0	0	0.00
30	T	Propionitrile	25.0000	23.6140	5.5	100	0.01
31	T	2,2-Dichloropropane	2.0000	2.1897	-9.5	100	0.00
32	T	cis-1,2-Dichloroethene	2.0000	2.0669	-3.3	100	0.00
33	C	Chloroform	2.0000	2.0616	-3.1	100	0.00
34	T	1-Bromopropane	2.0000	2.2551	-12.8	100	0.00
35	T	Bromochloromethane	2.0000	2.0347	-1.7	100	0.00
36	T	Tetrahydrofuran	25.0000	24.1040	3.6	100	0.00
37	S	Dibromofluoromethane	1.0000	0.9175	8.3	100	0.00
38	T	1,1,1-Trichloroethane	2.0000	2.0696	-3.5	100	-0.01
39	T	Cyclohexane	2.0000	2.0442	-2.2	100	0.00
40	T	1,1-Dichloropropene	2.0000	1.9713	1.4	100	0.00
41	T	Carbon Tetrachloride	2.0000	2.1748	-8.7	100	0.00
42	T	Tert-Amyl-Methyl ether	25.0000	25.0128	-0.1	100	0.00
43	S	1,2-Dichloroethane-d4	1.0000	0.9755	2.4	100	0.00
44	T	1,2-Dichloroethane	2.0000	2.0689	-3.4	100	0.00
45	T	Benzene	2.0000	2.1160	-5.8	100	0.00
46	T	Trichloroethene	2.0000	2.1089	-5.4	100	-0.01
47	T	Methylcyclohexane	2.0000	2.0212	-1.1	100	0.00
48	C	1,2-Dichloropropane	2.0000	2.0525	-2.6	100	0.00
49	T	1,4-Dioxane	50.0000	49.5313	0.9	100	0.00
50	T	Bromodichloromethane	2.0000	1.9996	0.0	100	0.00
51	T	Dibromomethane	2.0000	1.9537	2.3	100	0.01
52	T	2-Chloroethyl Vinyl Ether	-1.0000	1.8512	0.0	100	0.00
53	T	4-Methyl-2-Pentanone	-1.0000	1.4573	0.0	100	0.00
54	T	cis-1,3-Dichloropropene	2.0000	1.9172	4.1	100	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\051316\11M11840.D Vial: 6
Acq On : 13 May 2016 16:15 Operator: JDS
Sample : WG568769-05 2.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	Dimethyl Disulfide	-1.0000	1.6278	0.0	100	0.00
56	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
57	S	Toluene-d8	1.0000	1.0138	-1.4	100	0.00
58	C	Toluene	2.0000	2.0332	-1.7	100	0.00
59	T	Ethyl Methacrylate	2.0000	1.7234	13.8	100	-0.01
60	T	trans-1,3-Dichloropropene	2.0000	1.9114	4.4	100	0.00
61	T	1,1,2-Trichloroethane	2.0000	2.1674	-8.4	100	0.00
62	T	2-Hexanone	-1.0000	1.7098	0.0	100	0.00
63	T	1,3-Dichloropropane	2.0000	1.9471	2.6	100	0.00
64	T	Tetrachloroethene	2.0000	2.1547	-7.7	100	0.00
65	T	Dibromochloromethane	2.0000	1.9620	1.9	100	0.00
66	T	1,2-Dibromoethane	2.0000	2.0238	-1.2	100	0.00
67	T	1-Chlorohexane	2.0000	1.9844	0.8	100	0.00
68	P	Chlorobenzene	2.0000	2.1520	-7.6	100	0.00
69	T	1,1,1,2-Tetrachloroethane	2.0000	2.1574	-7.9	100	0.00
70	C	Ethylbenzene	2.0000	2.1518	-7.6	100	0.00
71	T	m-,p-Xylene	4.0000	4.1851	-4.6	100	0.00
72	T	o-Xylene	2.0000	2.0398	-2.0	100	0.00
73	T	Styrene	2.0000	1.9307	3.5	100	0.01
74	P	Bromoform	2.0000	1.8176	9.1	100	0.00
75	T	Isopropylbenzene	2.0000	2.1063	-5.3	100	0.00
76	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
77	P	1,1,2,2-Tetrachloroethane	2.0000	1.8958	5.2	100	-0.01
78	S	p-Bromofluorobenzene	1.0000	1.0013	-0.1	100	0.00
79	T	1,2,3-Trichloropropane	2.0000	2.1172	-5.9	100	0.00
80	T	trans-1,4-Dichloro-2-Butene	2.0000	1.9762	1.2	100	0.00
81	T	n-Propylbenzene	2.0000	2.1256	-6.3	100	0.00
82	T	Bromobenzene	2.0000	2.1009	-5.0	100	-0.01
83	T	1,3,5-Trimethylbenzene	2.0000	2.0481	-2.4	100	-0.01
84	T	2-Chlorotoluene	2.0000	2.0925	-4.6	100	0.00
85	T	4-Chlorotoluene	2.0000	2.1326	-6.6	100	0.00
86	T	a-Methylstyrene	2.0000	1.8297	8.5	100	0.00
87	T	tert-Butylbenzene	2.0000	2.0497	-2.5	100	0.00
88	T	1,2,4-Trimethylbenzene	2.0000	2.0213	-1.1	100	0.00
89	T	sec-Butylbenzene	2.0000	2.0611	-3.1	100	0.00
90	T	p-Isopropyltoluene	2.0000	2.1272	-6.4	100	0.00
91	T	1,3-Dichlorobenzene	2.0000	2.0970	-4.8	100	0.00
92	T	1,4-Dichlorobenzene	2.0000	2.1095	-5.5	100	0.00
93	T	n-Butylbenzene	2.0000	2.1373	-6.9	100	0.00
94	T	1,2-Dichlorobenzene	2.0000	2.1425	-7.1	100	0.00
95	T	1,2-Dibromo-3-Chloropropane	2.0000	1.9175	4.1	100	0.00
96	T	1,2,4-Trichlorobenzene	2.0000	1.9565	2.2	100	0.00
97	T	Hexachlorobutadiene	2.0000	2.0914	-4.6	100	0.00
98	T	Naphthalene	2.0000	1.9443	2.8	100	0.00
99	T	1,2,3-Trichlorobenzene	2.0000	2.0449	-2.2	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
11M11840.D 8260WT.M Sat May 14 18:46:53 2016 Page 2

Quantitation Report (Qedit)

Data File : C:\MSDCHEM\1\DATA\051316\11M11840.D Vial: 6
Acq On : 13 May 2016 16:15 Operator: JDS
Sample : WG568769-05 2.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:46 2016 Quant Results File: temp.res

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Single Level Calibration





TIC: 11M11840.D

(18) Methyl acetate (T)

7.01min 3.69ug/L

response 17794

Ion	Exp%	Act%
43.00	100	100
74.00	13.60	2.82#
59.00	6.50	1.21#
0.00	0.00	0.00

11M11840.D 8260WT.M Sat May 14 18:46:34 2016

Quantitation Report (Qedit)

Data File : C:\MSDCHEM\1\DATA\051316\11M11840.D Vial: 6
Acq On : 13 May 2016 16:15 Operator: JDS
Sample : WG568769-05 2.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:46 2016 Quant Results File: temp.res

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Single Level Calibration





TIC: 11M11840.D

(18) Methyl acetate (T)

7.06min 0.86ug/L mint

response 7344

Ion	Exp%	Act%
43.00	100	100
74.00	13.60	6.84#
59.00	6.50	2.93#
0.00	0.00	0.00

11M11840.D 8260WT.M Sat May 14 18:46:39 2016

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11841.D Vial: 7
Acq On : 13 May 2016 16:47 Operator: JDS
Sample : WG568769-06 5.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:12:58 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 17:07:56 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	510485	25.00 ug/L	-0.01
56) Chlorobenzene-d5	14.25	117	424113	25.00 ug/L	-0.01
76) 1,4-Dichlorobenzene-d4	17.06	152	244127	25.00 ug/L	-0.01

System Monitoring Compounds					
37) Dibromofluoromethane	9.63	111	13577	2.3093 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery =	9.24%#	
43) 1,2-Dichloroethane-d4	10.23	65	16340	2.3640 ug/L	-0.01
Spiked Amount 25.000	Range 80 - 120		Recovery =	9.44%#	
57) Toluene-d8	12.47	98	46369	2.3570 ug/L	-0.01
Spiked Amount 25.000	Range 88 - 110		Recovery =	9.44%#	
78) p-Bromofluorobenzene	15.64	95	19139	2.4218 ug/L	-0.01
Spiked Amount 25.000	Range 86 - 115		Recovery =	9.68%#	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	33260	4.2909 ug/L	97
3) Chloromethane	3.72	50	29484	4.2588 ug/L	100
4) Vinyl Chloride	3.96	62	24947	4.6446 ug/L	98
5) 1,3-Butadiene	4.00	54	24704	4.8274 ug/L	87
6) Bromomethane	4.85	94	16794	4.7876 ug/L	96
7) Chloroethane	5.00	64	15417	4.1982 ug/L	96
8) Trichlorofluoromethane	5.48	101	50663	4.8750 ug/L	98
9) Diethyl ether	6.01	59	199353	45.0064 ug/L	100
10) Isoprene	6.04	67	30513	4.3864 ug/L	99
11) Acrolein	6.25	56	12345	23.3877 ug/L	87
12) 1,1,2-Trichloro-1,2,2-Trif	6.26	101	23865	4.5585 ug/L	97
13) Acetone	6.35	43	6439	4.2186 ug/L	86
14) 1,1-Dichloroethene	6.56	61	43769	4.6158 ug/L	96
15) Tert-Butyl Alcohol	6.66	59	34212	87.4805 ug/L	100
16) Dimethyl Sulfide	6.81	62	17251	4.0482 ug/L	99
17) Iodomethane	7.06	142	7385	2.0684 ug/L	94
18) Methyl acetate	7.06	43	32363	5.4250 ug/L #	83
19) Methylene Chloride	7.31	84	23476	4.4707 ug/L	98
20) Carbon Disulfide	7.36	76	72931	4.4715 ug/L	98
21) Acrylonitrile	7.49	53	38039	20.0847 ug/L	96
22) Methyl Tert Butyl Ether	7.52	73	57597	4.3071 ug/L	98
23) trans-1,2-Dichloroethene	7.74	96	23369	4.3390 ug/L	99
24) n-Hexane	7.82	57	38507	4.3615 ug/L	99
25) Diisopropyl ether	8.15	45	1105208	48.2180 ug/L	99
26) Vinyl Acetate	8.31	43	24007	7.3442 ug/L	97
27) 1,1-Dichloroethane	8.34	63	51535	4.8316 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	899266	48.1312 ug/L	99
29) 2-Butanone	8.88	43	8098	3.5469 ug/L	98
30) Propionitrile	8.98	54	26204	41.2561 ug/L	97
31) 2,2-Dichloropropane	9.09	77	40285	5.2526 ug/L	97
32) cis-1,2-Dichloroethene	9.15	96	27456	4.6081 ug/L	91
33) Chloroform	9.35	83	48958	4.7306 ug/L	97
34) 1-Bromopropane	9.47	122	4632	4.9482 ug/L	88
35) Bromochloromethane	9.57	130	17073	4.6807 ug/L	99
36) Tetrahydrofuran	9.60	42	59638	39.6839 ug/L	99
38) 1,1,1-Trichloroethane	9.85	97	48123	4.8983 ug/L	100
39) Cyclohexane	9.88	56	51945	4.5931 ug/L	98
40) 1,1-Dichloropropene	10.04	75	33867	4.6344 ug/L	99
41) Carbon Tetrachloride	10.17	117	47434	5.0851 ug/L	100
42) Tert-Amyl-Methyl ether	10.13	73	615710	46.3601 ug/L	99

(#) = qualifier out of range (m) = manual integration
11M11841.D 8260WT.M Fri May 13 17:12:59 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11841.D Vial: 7
Acq On : 13 May 2016 16:47 Operator: JDS
Sample : WG568769-06 5.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:12:58 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 17:07:56 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit	Qvalue
44)	1,2-Dichloroethane	10.34	62	40281	4.6727 ug/L	99
45)	Benzene	10.38	78	94223	4.6151 ug/L	99
46)	Trichloroethene	11.08	130	31195	4.7740 ug/L	96
47)	Methylcyclohexane	11.17	83	35731	4.4100 ug/L	96
48)	1,2-Dichloropropane	11.29	63	25973	4.4921 ug/L	93
49)	1,4-Dioxane	11.56	88	2747	80.2363 ug/L	93
50)	Bromodichloromethane	11.57	83	36997	4.7467 ug/L	99
51)	Dibromomethane	11.65	93	13963	4.5774 ug/L	97
52)	2-Chloroethyl Vinyl Ether	11.84	63	11417	3.7313 ug/L	95
53)	4-Methyl-2-Pentanone	11.87	58	6133	3.2690 ug/L	90
54)	cis-1,3-Dichloropropene	12.17	75	36461	4.5343 ug/L	100
55)	Dimethyl Disulfide	12.42	79	18914	3.7032 ug/L	99
58)	Toluene	12.56	91	107617	4.9467 ug/L	100
59)	Ethyl Methacrylate	12.65	69	21187	4.1021 ug/L	99
60)	trans-1,3-Dichloropropene	12.73	75	32824	4.5778 ug/L	99
61)	1,1,2-Trichloroethane	12.94	97	17791	4.3060 ug/L	98
62)	2-Hexanone	12.86	43	11606	3.4244 ug/L	99
63)	1,3-Dichloropropane	13.21	76	30121	4.5460 ug/L	97
64)	Tetrachloroethene	13.33	164	24241	4.8819 ug/L	95
65)	Dibromochloromethane	13.59	129	27989	4.7024 ug/L	95
66)	1,2-Dibromoethane	13.82	107	18147	4.4382 ug/L	96
67)	1-Chlorohexane	13.89	91	33195	4.6591 ug/L	97
68)	Chlorobenzene	14.29	112	79059	4.9583 ug/L	98
69)	1,1,1,2-Tetrachloroethane	14.32	131	31047	4.8956 ug/L	97
70)	Ethylbenzene	14.31	106	39304	4.8168 ug/L	95
71)	m-,p-Xylene	14.39	106	95751	9.7273 ug/L	95
72)	o-Xylene	14.92	106	45759	4.7228 ug/L	96
73)	Styrene	14.95	104	77900	4.8227 ug/L	98
74)	Bromoform	15.43	173	15663	4.1856 ug/L	97
75)	Isopropylbenzene	15.31	105	123883	4.9827 ug/L	99
77)	1,1,2,2-Tetrachloroethane	15.51	83	18105	4.4049 ug/L	100
79)	1,2,3-Trichloropropane	15.70	110	6717	4.7266 ug/L	85
80)	trans-1,4-Dichloro-2-Buten	15.74	53	6067	3.6507 ug/L	69
81)	n-Propylbenzene	15.79	91	143972	5.2435 ug/L	99
82)	Bromobenzene	15.92	156	37345	4.9800 ug/L	93
83)	1,3,5-Trimethylbenzene	15.95	105	106030	5.0309 ug/L	100
84)	2-Chlorotoluene	16.05	91	99279	4.9734 ug/L	100
85)	4-Chlorotoluene	16.09	91	88275	5.1992 ug/L	98
86)	a-Methylstyrene	16.34	118	53766	4.7153 ug/L	98
87)	tert-Butylbenzene	16.39	134	23868	5.2956 ug/L	97
88)	1,2,4-Trimethylbenzene	16.44	105	112072	5.2401 ug/L	99
89)	sec-Butylbenzene	16.65	105	129249	5.1810 ug/L	99
90)	p-Isopropyltoluene	16.79	119	119605	5.2524 ug/L	98
91)	1,3-Dichlorobenzene	16.98	146	71265	5.0271 ug/L	99
92)	1,4-Dichlorobenzene	17.10	146	72437	5.0122 ug/L	96
93)	n-Butylbenzene	17.28	91	100062	4.9277 ug/L	99
94)	1,2-Dichlorobenzene	17.57	146	64925	4.9520 ug/L	98
95)	1,2-Dibromo-3-Chloropropan	18.49	75	3875	4.6567 ug/L	87
96)	1,2,4-Trichlorobenzene	19.55	180	46664	4.8629 ug/L	97
97)	Hexachlorobutadiene	19.69	225	20419	5.1319 ug/L	96
98)	Naphthalene	19.90	128	83612	4.4869 ug/L	99
99)	1,2,3-Trichlorobenzene	20.19	180	40714	4.5323 ug/L	98

(#) = qualifier out of range (m) = manual integration
11M11841.D 8260WT.M Fri May 13 17:12:59 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11841.D Vial: 7
Acq On : 13 May 2016 16:47 Operator: JDS
Sample : WG568769-06 5.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 13 17:12 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Fri May 13 17:07:56 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\051316\11M11841.D Vial: 7
Acq On : 13 May 2016 16:47 Operator: JDS
Sample : WG568769-06 5.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	100	0.00
2	T	Dichlorodifluoromethane	5.0000	4.1584	16.8	100	0.00
3	P	Chloromethane	5.0000	4.6027	7.9	100	0.00
4	C	Vinyl Chloride	5.0000	4.6895	6.2	100	0.00
5	T	1,3-Butadiene	5.0000	4.6702	6.6	100	0.00
6	T	Bromomethane	5.0000	4.5915	8.2	100	0.00
7	T	Chloroethane	5.0000	4.5105	9.8	100	0.00
8	T	Trichlorofluoromethane	5.0000	4.7781	4.4	100	0.00
9	T	Diethyl ether	50.0000	50.8371	-1.7	100	0.00
10	T	Isoprene	5.0000	4.6688	6.6	100	0.00
11	T	Acrolein	25.0000	24.7297	1.1	100	0.01
12	T	1,1,2-Trichloro-1,2,2-Trifl	5.0000	4.6385	7.2	100	0.01
13	T	Acetone	5.0000	5.3120	-6.2	100	0.01
14	C	1,1-Dichloroethene	5.0000	4.6835	6.3	100	0.00
15	T	Tert-Butyl Alcohol	100.0000	102.0360	-2.0	100	0.00
16	T	Dimethyl Sulfide	5.0000	4.6223	7.6	100	0.00
17	T	Iodomethane	5.0000	3.9124	21.8	100	0.00
18	T	Methyl acetate	5.0000	7.5474	-50.9#	100	0.00
19	T	Methylene Chloride	5.0000	4.8539	2.9	100	0.00
20	T	Carbon Disulfide	5.0000	4.6615	6.8	100	0.00
21	T	Acrylonitrile	25.0000	23.4049	6.4	100	0.00
22	T	Methyl Tert Butyl Ether	5.0000	4.7747	4.5	100	0.00
23	T	trans-1,2-Dichloroethene	5.0000	4.5182	9.6	100	0.00
24	T	n-Hexane	5.0000	4.4700	10.6	100	0.00
25	T	Diisopropyl ether	50.0000	51.4154	-2.8	100	0.00
26	T	Vinyl Acetate	5.0000	5.1698	-3.4	100	0.00
27	P	1,1-Dichloroethane	5.0000	4.9658	0.7	100	0.00
28	T	Ethyl-Tert-Butyl ether	50.0000	51.3996	-2.8	100	0.00
29	T	2-Butanone	5.0000	4.4190	11.6	100	0.01
30	T	Propionitrile	50.0000	48.8210	2.4	100	0.00
31	T	2,2-Dichloropropane	5.0000	4.9614	0.8	100	0.00
32	T	cis-1,2-Dichloroethene	5.0000	4.8145	3.7	100	0.00
33	C	Chloroform	5.0000	4.8039	3.9	100	0.00
34	T	1-Bromopropane	5.0000	5.1449	-2.9	100	0.00
35	T	Bromochloromethane	5.0000	5.0410	-0.8	100	0.00
36	T	Tetrahydrofuran	50.0000	46.0861	7.8	100	0.00
37	S	Dibromofluoromethane	2.5000	2.4302	2.8	100	0.00
38	T	1,1,1-Trichloroethane	5.0000	4.7984	4.0	100	0.00
39	T	Cyclohexane	5.0000	4.7526	4.9	100	0.00
40	T	1,1-Dichloropropene	5.0000	4.7295	5.4	100	0.00
41	T	Carbon Tetrachloride	5.0000	4.8435	3.1	100	-0.01
42	T	Tert-Amyl-Methyl ether	50.0000	50.6462	-1.3	100	0.00
43	S	1,2-Dichloroethane-d4	2.5000	2.5650	-2.6	100	0.00
44	T	1,2-Dichloroethane	5.0000	4.8665	2.7	100	0.00
45	T	Benzene	5.0000	4.8065	3.9	100	0.00
46	T	Trichloroethene	5.0000	4.8959	2.1	100	-0.01
47	T	Methylcyclohexane	5.0000	4.5439	9.1	100	0.00
48	C	1,2-Dichloropropane	5.0000	4.7864	4.3	100	0.00
49	T	1,4-Dioxane	100.0000	102.6541	-2.7	100	0.01
50	T	Bromodichloromethane	5.0000	4.8566	2.9	100	0.00
51	T	Dibromomethane	5.0000	4.9379	1.2	100	0.00
52	T	2-Chloroethyl Vinyl Ether	5.0000	4.2689	14.6	100	0.00
53	T	4-Methyl-2-Pentanone	5.0000	4.0658	18.7	100	0.00
54	T	cis-1,3-Dichloropropene	5.0000	4.7853	4.3	100	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\051316\11M11841.D Vial: 7
Acq On : 13 May 2016 16:47 Operator: JDS
Sample : WG568769-06 5.0ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	Dimethyl Disulfide	5.0000	4.1252	17.5	100	0.00
56	I	Chlorobenzene-d5	25.0000	25.0000	0.0	100	0.00
57	S	Toluene-d8	2.5000	2.5163	-0.7	100	0.00
58	C	Toluene	5.0000	5.0124	-0.2	100	0.00
59	T	Ethyl Methacrylate	5.0000	4.6214	7.6	100	0.00
60	T	trans-1,3-Dichloropropene	5.0000	4.7629	4.7	100	0.00
61	T	1,1,2-Trichloroethane	5.0000	4.6730	6.5	100	0.00
62	T	2-Hexanone	5.0000	4.1507	17.0	100	0.00
63	T	1,3-Dichloropropane	5.0000	4.9267	1.5	100	0.00
64	T	Tetrachloroethene	5.0000	4.8551	2.9	100	-0.01
65	T	Dibromochloromethane	5.0000	4.8199	3.6	100	0.00
66	T	1,2-Dibromoethane	5.0000	4.7575	4.8	100	0.00
67	T	1-Chlorohexane	5.0000	4.6895	6.2	100	0.00
68	P	Chlorobenzene	5.0000	5.0264	-0.5	100	0.00
69	T	1,1,1,2-Tetrachloroethane	5.0000	4.8783	2.4	100	0.00
70	C	Ethylbenzene	5.0000	4.8732	2.5	100	0.00
71	T	m-,p-Xylene	10.0000	9.7654	2.3	100	0.00
72	T	o-Xylene	5.0000	4.7954	4.1	100	0.00
73	T	Styrene	5.0000	4.9411	1.2	100	0.00
74	P	Bromoform	5.0000	4.4503	11.0	100	0.00
75	T	Isopropylbenzene	5.0000	4.9091	1.8	100	0.00
76	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	100	0.00
77	P	1,1,2,2-Tetrachloroethane	5.0000	4.7624	4.8	100	-0.01
78	S	p-Bromofluorobenzene	2.5000	2.5432	-1.7	100	0.00
79	T	1,2,3-Trichloropropane	5.0000	5.0393	-0.8	100	0.00
80	T	trans-1,4-Dichloro-2-Butene	5.0000	4.0050	19.9	100	0.00
81	T	n-Propylbenzene	5.0000	5.0923	-1.8	100	0.00
82	T	Bromobenzene	5.0000	5.0338	-0.7	100	0.00
83	T	1,3,5-Trimethylbenzene	5.0000	4.9145	1.7	100	-0.01
84	T	2-Chlorotoluene	5.0000	5.0134	-0.3	100	0.00
85	T	4-Chlorotoluene	5.0000	4.9765	0.5	100	0.00
86	T	a-Methylstyrene	5.0000	4.7239	5.5	100	0.00
87	T	tert-Butylbenzene	5.0000	5.2217	-4.4	100	-0.01
88	T	1,2,4-Trimethylbenzene	5.0000	5.1012	-2.0	100	0.00
89	T	sec-Butylbenzene	5.0000	4.9892	0.2	100	0.00
90	T	p-Isopropyltoluene	5.0000	5.0265	-0.5	100	0.00
91	T	1,3-Dichlorobenzene	5.0000	4.9737	0.5	100	0.00
92	T	1,4-Dichlorobenzene	5.0000	4.9875	0.2	100	0.00
93	T	n-Butylbenzene	5.0000	4.7722	4.6	100	0.00
94	T	1,2-Dichlorobenzene	5.0000	4.9741	0.5	100	0.00
95	T	1,2-Dibromo-3-Chloropropane	5.0000	5.1051	-2.1	100	0.00
96	T	1,2,4-Trichlorobenzene	5.0000	4.8858	2.3	100	0.00
97	T	Hexachlorobutadiene	5.0000	4.8528	2.9	100	0.00
98	T	Naphthalene	5.0000	4.7164	5.7	100	0.00
99	T	1,2,3-Trichlorobenzene	5.0000	4.6448	7.1	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0
11M11841.D 8260WT.M Sat May 14 18:48:20 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11842.D Vial: 7
Acq On : 13 May 2016 17:19 Operator: JDS
Sample : WG568769-07 20ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50:53 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	512017	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	433937	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	255524	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	58089	10.3664 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 41.48%#		
43) 1,2-Dichloroethane-d4	10.23	65	66804	10.4555 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 41.84%#		
57) Toluene-d8	12.47	98	193532	10.2645 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 41.04%#		
78) p-Bromofluorobenzene	15.64	95	81186	10.3070 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 41.24%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	176636	22.0179 ug/L	99
3) Chloromethane	3.72	50	126901	19.7510 ug/L	98
4) Vinyl Chloride	3.96	62	109592	20.5394 ug/L	98
5) 1,3-Butadiene	4.00	54	116317	21.9237 ug/L	96
6) Bromomethane	4.85	94	73878	20.1380 ug/L	93
7) Chloroethane	5.00	64	71103	20.7401 ug/L	100
8) Trichlorofluoromethane	5.49	101	225913	21.2425 ug/L	99
9) Diethyl ether	6.01	59	321203	81.6650 ug/L	99
10) Isoprene	6.04	67	135641	20.6925 ug/L	98
11) Acrolein	6.24	56	19366	37.8284 ug/L	95
12) 1,1,2-Trichloro-1,2,2-Trif	6.25	101	109792	21.2757 ug/L	99
13) Acetone	6.34	43	27399	22.5356 ug/L	96
14) 1,1-Dichloroethene	6.56	61	201658	21.5139 ug/L	98
15) Tert-Butyl Alcohol	6.67	59	60081	178.6532 ug/L	99
16) Dimethyl Sulfide	6.81	62	75759	20.2383 ug/L	97
17) Iodomethane	7.06	142	67280	17.2594 ug/L	97
18) Methyl acetate	7.07	43	87945	22.4983 ug/L	96
19) Methylene Chloride	7.31	84	99453	20.5013 ug/L	100
20) Carbon Disulfide	7.36	76	327043	20.8410 ug/L	100
21) Acrylonitrile	7.49	53	69618	42.7070 ug/L	95
22) Methyl Tert Butyl Ether	7.52	73	266288	22.0086 ug/L	99
23) trans-1,2-Dichloroethene	7.75	96	107858	20.7912 ug/L	99
24) n-Hexane	7.82	57	184135	21.3111 ug/L	99
25) Diisopropyl ether	8.15	45	1729257	80.2062 ug/L	99
26) Vinyl Acetate	8.31	43	160029	20.6235 ug/L	99
27) 1,1-Dichloroethane	8.34	63	218994	21.0387 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	1435318	81.7933 ug/L	100
29) 2-Butanone	8.87	43	39878	21.6958 ug/L	99
30) Propionitrile	8.97	54	47221	87.7149 ug/L	99
31) 2,2-Dichloropropane	9.09	77	166427	20.4352 ug/L	100
32) cis-1,2-Dichloroethene	9.15	96	120620	21.0877 ug/L	99
33) Chloroform	9.35	83	208615	20.4085 ug/L	99
34) 1-Bromopropane	9.48	122	19153	19.2424 ug/L	96
35) Bromochloromethane	9.57	130	77189	22.7226 ug/L	99
36) Tetrahydrofuran	9.60	42	109362	84.2581 ug/L	100
38) 1,1,1-Trichloroethane	9.85	97	210871	20.9631 ug/L	98
39) Cyclohexane	9.88	56	235071	21.4428 ug/L	99
40) 1,1-Dichloropropene	10.04	75	151682	21.1188 ug/L	99
41) Carbon Tetrachloride	10.17	117	212111	21.5941 ug/L	100
42) Tert-Amyl-Methyl ether	10.13	73	1001790	82.1572 ug/L	100

(#) = qualifier out of range (m) = manual integration
11M11842.D 8260WT.M Sat May 14 18:50:54 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11842.D Vial: 7
Acq On : 13 May 2016 17:19 Operator: JDS
Sample : WG568769-07 20ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50:53 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
44)	1,2-Dichloroethane	10.34	62	180438	21.7339	ug/L	99
45)	Benzene	10.38	78	407777	20.7394	ug/L	100
46)	Trichloroethene	11.08	130	132094	20.6692	ug/L	99
47)	Methylcyclohexane	11.17	83	168355	21.3455	ug/L	98
48)	1,2-Dichloropropane	11.29	63	112875	20.7388	ug/L	99
49)	1,4-Dioxane	11.56	88	4948	165.8209	ug/L	99
50)	Bromodichloromethane	11.57	83	163762	21.4329	ug/L	98
51)	Dibromomethane	11.65	93	61028	21.5174	ug/L	98
52)	2-Chloroethyl Vinyl Ether	11.84	63	56664	21.1238	ug/L	99
53)	4-Methyl-2-Pentanone	11.87	58	32621	21.5609	ug/L	99
54)	cis-1,3-Dichloropropene	12.17	75	166399	21.7735	ug/L	100
55)	Dimethyl Disulfide	12.42	79	93808	20.3984	ug/L	99
58)	Toluene	12.56	91	467769	21.2936	ug/L	99
59)	Ethyl Methacrylate	12.64	69	103892	22.1484	ug/L	98
60)	trans-1,3-Dichloropropene	12.73	75	155661	22.0759	ug/L	100
61)	1,1,2-Trichloroethane	12.94	97	85297	21.8971	ug/L	100
62)	2-Hexanone	12.86	43	62113	21.7109	ug/L	99
63)	1,3-Dichloropropane	13.21	76	137448	21.9726	ug/L	98
64)	Tetrachloroethene	13.34	164	106368	20.8217	ug/L	99
65)	Dibromochloromethane	13.59	129	131143	22.0723	ug/L	99
66)	1,2-Dibromoethane	13.82	107	84381	21.6209	ug/L	99
67)	1-Chlorohexane	13.89	91	153420	21.1830	ug/L	100
68)	Chlorobenzene	14.29	112	338669	21.0446	ug/L	99
69)	1,1,1,2-Tetrachloroethane	14.32	131	134571	20.6660	ug/L	98
70)	Ethylbenzene	14.31	106	171752	20.8130	ug/L	97
71)	m-,p-Xylene	14.39	106	419228	41.7882	ug/L	99
72)	o-Xylene	14.92	106	208125	21.3170	ug/L	99
73)	Styrene	14.95	104	353452	21.9116	ug/L	98
74)	Bromoform	15.43	173	78592	21.8247	ug/L	100
75)	Isopropylbenzene	15.31	105	552747	21.4078	ug/L	99
77)	1,1,2,2-Tetrachloroethane	15.52	83	90122	22.6486	ug/L	98
79)	1,2,3-Trichloropropane	15.70	110	30969	20.9562	ug/L	97
80)	trans-1,4-Dichloro-2-Buten	15.74	53	36190	19.3325	ug/L	86
81)	n-Propylbenzene	15.79	91	640711	21.6512	ug/L	100
82)	Bromobenzene	15.92	156	160227	20.6340	ug/L	100
83)	1,3,5-Trimethylbenzene	15.95	105	481759	21.3337	ug/L	100
84)	2-Chlorotoluene	16.05	91	435100	20.9919	ug/L	99
85)	4-Chlorotoluene	16.09	91	391759	21.1004	ug/L	100
86)	a-Methylstyrene	16.34	118	253642	21.2911	ug/L	99
87)	tert-Butylbenzene	16.39	134	101859	21.2903	ug/L	99
88)	1,2,4-Trimethylbenzene	16.44	105	505239	21.9714	ug/L	98
89)	sec-Butylbenzene	16.65	105	586770	21.6398	ug/L	100
90)	p-Isopropyltoluene	16.79	119	539795	21.6735	ug/L	100
91)	1,3-Dichlorobenzene	16.98	146	312832	20.8591	ug/L	99
92)	1,4-Dichlorobenzene	17.10	146	314339	20.6779	ug/L	99
93)	n-Butylbenzene	17.28	91	474312	21.6121	ug/L	99
94)	1,2-Dichlorobenzene	17.57	146	287316	21.0305	ug/L	100
95)	1,2-Dibromo-3-Chloropropan	18.49	75	19607	22.2193	ug/L	89
96)	1,2,4-Trichlorobenzene	19.55	180	212916	21.2982	ug/L	100
97)	Hexachlorobutadiene	19.69	225	91388	20.7507	ug/L	99
98)	Naphthalene	19.90	128	414073	22.3154	ug/L	99
99)	1,2,3-Trichlorobenzene	20.19	180	193437	21.0836	ug/L	99

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11842.D Vial: 7
Acq On : 13 May 2016 17:19 Operator: JDS
Sample : WG568769-07 20ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11843.D Vial: 8
Acq On : 13 May 2016 17:51 Operator: JDS
Sample : WG568769-08 50ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50:56 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	521583	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	438582	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	263885	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.63	111	139784	24.4879 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 97.96%		
43) 1,2-Dichloroethane-d4	10.23	65	158200	24.3058 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 97.24%		
57) Toluene-d8	12.47	98	462969	24.2947 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 97.16%		
78) p-Bromofluorobenzene	15.64	95	188875	23.2189 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 92.88%		
Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	416971	51.0228 ug/L	100
3) Chloromethane	3.72	50	287508	43.9273 ug/L	100
4) Vinyl Chloride	3.96	62	261024	48.0232 ug/L	100
5) 1,3-Butadiene	4.00	54	280913	51.9760 ug/L	100
6) Bromomethane	4.85	94	172294	46.1032 ug/L	100
7) Chloroethane	5.00	64	167978	48.0991 ug/L	100
8) Trichlorofluoromethane	5.48	101	519806	47.9806 ug/L	100
9) Diethyl ether	6.01	59	391891	97.8098 ug/L	100
10) Isoprene	6.04	67	331169	49.5943 ug/L	100
11) Acrolein	6.23	56	25228	47.9550 ug/L	100
12) 1,1,2-Trichloro-1,2,2-Trif	6.25	101	255341	48.5730 ug/L	100
13) Acetone	6.34	43	57896	46.7459 ug/L	98
14) 1,1-Dichloroethene	6.56	61	459793	48.1534 ug/L	100
15) Tert-Butyl Alcohol	6.66	59	69395	202.5643 ug/L	100
16) Dimethyl Sulfide	6.81	62	193959	50.8640 ug/L	100
17) Iodomethane	7.06	142	214554	49.3519 ug/L	100
18) Methyl acetate	7.06	43	185465	48.1983 ug/L	100
19) Methylene Chloride	7.31	84	233601	47.2715 ug/L	100
20) Carbon Disulfide	7.36	76	816514	51.0786 ug/L	100
21) Acrylonitrile	7.49	53	84821	51.0789 ug/L	100
22) Methyl Tert Butyl Ether	7.52	73	610666	49.5457 ug/L	100
23) trans-1,2-Dichloroethene	7.74	96	252405	47.7624 ug/L	100
24) n-Hexane	7.82	57	449694	51.0913 ug/L	100
25) Diisopropyl ether	8.15	45	2181723	99.3365 ug/L	100
26) Vinyl Acetate	8.31	43	410275	48.2141 ug/L	100
27) 1,1-Dichloroethane	8.34	63	516691	48.7281 ug/L	100
28) Ethyl-Tert-Butyl ether	8.70	59	1783350	99.7625 ug/L	100
29) 2-Butanone	8.87	43	91238	48.7281 ug/L	100
30) Propionitrile	8.97	54	56538	103.0954 ug/L	100
31) 2,2-Dichloropropane	9.09	77	402511	48.5170 ug/L	100
32) cis-1,2-Dichloroethene	9.15	96	282222	48.4353 ug/L	100
33) Chloroform	9.35	83	489993	47.0562 ug/L	100
34) 1-Bromopropane	9.48	122	50180	48.4991 ug/L	100
35) Bromochloromethane	9.57	130	179921	51.9932 ug/L	100
36) Tetrahydrofuran	9.60	42	127999	96.8083 ug/L	100
38) 1,1,1-Trichloroethane	9.85	97	497667	48.5667 ug/L	100
39) Cyclohexane	9.88	56	575045	51.4928 ug/L	100
40) 1,1-Dichloropropene	10.04	75	353148	48.2674 ug/L	100
41) Carbon Tetrachloride	10.18	117	493225	49.2921 ug/L	100
42) Tert-Amyl-Methyl ether	10.13	73	1227500	98.8215 ug/L	100

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11843.D Vial: 8
Acq On : 13 May 2016 17:51 Operator: JDS
Sample : WG568769-08 50ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50:56 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
44) 1,2-Dichloroethane	10.34	62	419000	49.5434	ug/L	100
45) Benzene	10.38	78	947514	47.3064	ug/L	100
46) Trichloroethene	11.09	130	307211	47.1888	ug/L	100
47) Methylcyclohexane	11.17	83	418631	52.1043	ug/L	100
48) 1,2-Dichloropropane	11.29	63	268423	48.4136	ug/L	100
49) 1,4-Dioxane	11.55	88	5866	189.1665	ug/L	100
50) Bromodichloromethane	11.57	83	385292	49.5015	ug/L	100
51) Dibromomethane	11.65	93	142910	49.4635	ug/L	100
52) 2-Chloroethyl Vinyl Ether	11.84	63	133225	48.7541	ug/L	100
53) 4-Methyl-2-Pentanone	11.87	58	75720	49.1295	ug/L	100
54) cis-1,3-Dichloropropene	12.17	75	394857	50.7199	ug/L	100
55) Dimethyl Disulfide	12.42	79	236984	50.5867	ug/L	100
58) Toluene	12.56	91	1095397	49.3360	ug/L	100
59) Ethyl Methacrylate	12.65	69	249821	52.6944	ug/L	100
60) trans-1,3-Dichloropropene	12.73	75	366443	51.4187	ug/L	100
61) 1,1,2-Trichloroethane	12.93	97	194150	49.3135	ug/L	100
62) 2-Hexanone	12.86	43	144364	49.9264	ug/L	100
63) 1,3-Dichloropropane	13.21	76	316783	50.1049	ug/L	100
64) Tetrachloroethene	13.34	164	251037	48.6205	ug/L	100
65) Dibromochloromethane	13.59	129	310662	51.7330	ug/L	100
66) 1,2-Dibromoethane	13.82	107	196704	49.8676	ug/L	100
67) 1-Chlorohexane	13.89	91	376690	51.4596	ug/L	100
68) Chlorobenzene	14.29	112	797442	49.0276	ug/L	100
69) 1,1,1,2-Tetrachloroethane	14.32	131	325801	49.5033	ug/L	100
70) Ethylbenzene	14.31	106	408749	49.0078	ug/L	100
71) m-,p-Xylene	14.39	106	996141	98.2427	ug/L	100
72) o-Xylene	14.92	106	492043	49.8632	ug/L	100
73) Styrene	14.95	104	838125	51.4077	ug/L	100
74) Bromoform	15.43	173	184100	50.5823	ug/L	100
75) Isopropylbenzene	15.31	105	1309383	50.1751	ug/L	100
77) 1,1,2,2-Tetrachloroethane	15.52	83	208930	50.8426	ug/L	100
79) 1,2,3-Trichloropropane	15.70	110	72411	46.9859	ug/L	100
80) trans-1,4-Dichloro-2-Buten	15.74	53	94075	47.5347	ug/L	100
81) n-Propylbenzene	15.79	91	1522917	49.8325	ug/L	100
82) Bromobenzene	15.92	156	379178	47.2834	ug/L	100
83) 1,3,5-Trimethylbenzene	15.96	105	1143246	49.0221	ug/L	100
84) 2-Chlorotoluene	16.05	91	1021921	47.7415	ug/L	100
85) 4-Chlorotoluene	16.09	91	930690	48.5394	ug/L	100
86) a-Methylstyrene	16.34	118	649847	52.8208	ug/L	100
87) tert-Butylbenzene	16.40	134	241955	48.9705	ug/L	100
88) 1,2,4-Trimethylbenzene	16.44	105	1178963	49.6454	ug/L	100
89) sec-Butylbenzene	16.65	105	1390388	49.6522	ug/L	100
90) p-Isopropyltoluene	16.79	119	1282252	49.8529	ug/L	100
91) 1,3-Dichlorobenzene	16.98	146	737967	47.6473	ug/L	100
92) 1,4-Dichlorobenzene	17.10	146	740977	47.1987	ug/L	100
93) n-Butylbenzene	17.28	91	1126170	49.6883	ug/L	100
94) 1,2-Dichlorobenzene	17.57	146	684016	48.4812	ug/L	100
95) 1,2-Dibromo-3-Chloropropan	18.49	75	42848	46.3023	ug/L	100
96) 1,2,4-Trichlorobenzene	19.55	180	512927	49.6830	ug/L	100
97) Hexachlorobutadiene	19.69	225	219129	48.1793	ug/L	100
98) Naphthalene	19.90	128	962852	50.2463	ug/L	100
99) 1,2,3-Trichlorobenzene	20.19	180	453828	47.8975	ug/L	100

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11843.D Vial: 8
Acq On : 13 May 2016 17:51 Operator: JDS
Sample : WG568769-08 50ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11844.D Vial: 9
Acq On : 13 May 2016 18:22 Operator: JDS
Sample : WG568769-09 100ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50:57 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	541540	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	458522	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	275719	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	307700	51.9175 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 207.68%#		
43) 1,2-Dichloroethane-d4	10.23	65	345779	51.1676 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 204.68%#		
57) Toluene-d8	12.47	98	1011334	50.7627 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 203.04%#		
78) p-Bromofluorobenzene	15.64	95	423157	49.7872 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 199.16%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	918088	108.2020 ug/L	100
3) Chloromethane	3.72	50	641242	94.3625 ug/L	99
4) Vinyl Chloride	3.95	62	586615	103.9481 ug/L	100
5) 1,3-Butadiene	3.99	54	571156	101.7838 ug/L	99
6) Bromomethane	4.85	94	394826	101.7560 ug/L	99
7) Chloroethane	5.00	64	370541	102.1913 ug/L	99
8) Trichlorofluoromethane	5.48	101	1135647	100.9627 ug/L	98
9) Diethyl ether	6.01	59	836495	201.0821 ug/L	99
10) Isoprene	6.04	67	699724	100.9257 ug/L	100
11) Acrolein	6.24	56	55228	99.4421 ug/L	99
12) 1,1,2-Trichloro-1,2,2-Trif	6.26	101	545013	99.8560 ug/L	100
13) Acetone	6.34	43	123799	96.2731 ug/L	95
14) 1,1-Dichloroethene	6.56	61	1004294	101.3022 ug/L	100
15) Tert-Butyl Alcohol	6.66	59	148134	416.4687 ug/L	98
16) Dimethyl Sulfide	6.81	62	401237	101.3432 ug/L	100
17) Iodomethane	7.06	142	471112	102.3553 ug/L	100
18) Methyl acetate	7.07	43	381031	98.0957 ug/L	99
19) Methylene Chloride	7.31	84	501609	97.7649 ug/L	99
20) Carbon Disulfide	7.36	76	1695322	102.1458 ug/L	100
21) Acrylonitrile	7.49	53	184461	106.9881 ug/L	99
22) Methyl Tert Butyl Ether	7.52	73	1312901	102.5951 ug/L	100
23) trans-1,2-Dichloroethene	7.74	96	542907	98.9478 ug/L	98
24) n-Hexane	7.82	57	921942	100.8849 ug/L	99
25) Diisopropyl ether	8.15	45	4599625	201.7086 ug/L	99
26) Vinyl Acetate	8.31	43	855488	94.3763 ug/L	99
27) 1,1-Dichloroethane	8.34	63	1111115	100.9254 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	3780769	203.7060 ug/L	100
29) 2-Butanone	8.87	43	198511	102.1130 ug/L	100
30) Propionitrile	8.98	54	118699	208.4678 ug/L	97
31) 2,2-Dichloropropane	9.09	77	914695	106.1904 ug/L	100
32) cis-1,2-Dichloroethene	9.15	96	606774	100.2977 ug/L	99
33) Chloroform	9.35	83	1053766	97.4685 ug/L	100
34) 1-Bromopropane	9.48	122	106914	98.8615 ug/L	100
35) Bromochloromethane	9.57	130	387239	107.7795 ug/L	100
36) Tetrahydrofuran	9.60	42	270148	196.7890 ug/L	99
38) 1,1,1-Trichloroethane	9.85	97	1088263	102.2884 ug/L	99
39) Cyclohexane	9.88	56	1213186	104.6320 ug/L	100
40) 1,1-Dichloropropene	10.04	75	775674	102.1101 ug/L	98
41) Carbon Tetrachloride	10.18	117	1074698	103.4456 ug/L	99
42) Tert-Amyl-Methyl ether	10.13	73	2619501	203.1147 ug/L	99

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11844.D Vial: 9
Acq On : 13 May 2016 18:22 Operator: JDS
Sample : WG568769-09 100ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50:57 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit	Qvalue
44)	1,2-Dichloroethane	10.34	62	895394	101.9714 ug/L	100
45)	Benzene	10.38	78	2041856	98.1867 ug/L	100
46)	Trichloroethene	11.08	130	677193	100.1861 ug/L	99
47)	Methylcyclohexane	11.17	83	867487	103.9915 ug/L	100
48)	1,2-Dichloropropane	11.29	63	580562	100.8530 ug/L	98
49)	1,4-Dioxane	11.56	88	13651	395.0901 ug/L	100
50)	Bromodichloromethane	11.57	83	837797	103.6717 ug/L	99
51)	Dibromomethane	11.65	93	315786	105.2707 ug/L	99
52)	2-Chloroethyl Vinyl Ether	11.84	63	293822	103.5626 ug/L	100
53)	4-Methyl-2-Pentanone	11.87	58	164890	103.0430 ug/L	98
54)	cis-1,3-Dichloropropene	12.17	75	854611	105.7305 ug/L	99
55)	Dimethyl Disulfide	12.42	79	505463	103.9202 ug/L	99
58)	Toluene	12.56	91	2361378	101.7300 ug/L	100
59)	Ethyl Methacrylate	12.65	69	529740	106.8782 ug/L	98
60)	trans-1,3-Dichloropropene	12.73	75	793063	106.4420 ug/L	99
61)	1,1,2-Trichloroethane	12.94	97	418327	101.6330 ug/L	99
62)	2-Hexanone	12.86	43	314588	104.0648 ug/L	99
63)	1,3-Dichloropropane	13.21	76	685844	103.7610 ug/L	100
64)	Tetrachloroethene	13.34	164	550335	101.9528 ug/L	99
65)	Dibromochloromethane	13.59	129	681591	108.5659 ug/L	100
66)	1,2-Dibromoethane	13.82	107	427945	103.7729 ug/L	99
67)	1-Chlorohexane	13.89	91	807177	105.4730 ug/L	98
68)	Chlorobenzene	14.29	112	1705646	100.3046 ug/L	99
69)	1,1,1,2-Tetrachloroethane	14.32	131	704959	102.4557 ug/L	100
70)	Ethylbenzene	14.31	106	883839	101.3614 ug/L	99
71)	m-,p-Xylene	14.39	106	2123651	200.3333 ug/L	98
72)	o-Xylene	14.92	106	1053682	102.1356 ug/L	100
73)	Styrene	14.95	104	1807222	106.0282 ug/L	99
74)	Bromoform	15.43	173	412378	108.3755 ug/L	100
75)	Isopropylbenzene	15.31	105	2813193	103.1127 ug/L	99
77)	1,1,2,2-Tetrachloroethane	15.52	83	454052	105.7501 ug/L	99
79)	1,2,3-Trichloropropane	15.70	110	156532	96.8210 ug/L	100
80)	trans-1,4-Dichloro-2-Buten	15.74	53	203327	97.5343 ug/L	96
81)	n-Propylbenzene	15.79	91	3285915	102.9061 ug/L	99
82)	Bromobenzene	15.92	156	828973	98.9359 ug/L	99
83)	1,3,5-Trimethylbenzene	15.96	105	2483230	101.9102 ug/L	100
84)	2-Chlorotoluene	16.05	91	2208423	98.7437 ug/L	100
85)	4-Chlorotoluene	16.09	91	2020827	100.8710 ug/L	100
86)	a-Methylstyrene	16.34	118	1361545	105.9190 ug/L	99
87)	tert-Butylbenzene	16.40	134	532522	103.1540 ug/L	98
88)	1,2,4-Trimethylbenzene	16.44	105	2546912	102.6457 ug/L	100
89)	sec-Butylbenzene	16.65	105	3003570	102.6569 ug/L	99
90)	p-Isopropyltoluene	16.79	119	2763313	102.8241 ug/L	99
91)	1,3-Dichlorobenzene	16.98	146	1600684	98.9134 ug/L	100
92)	1,4-Dichlorobenzene	17.10	146	1605475	97.8762 ug/L	100
93)	n-Butylbenzene	17.28	91	2452438	103.5609 ug/L	100
94)	1,2-Dichlorobenzene	17.57	146	1481276	100.4827 ug/L	100
95)	1,2-Dibromo-3-Chloropropan	18.49	75	95062	97.5959 ug/L	99
96)	1,2,4-Trichlorobenzene	19.55	180	1100416	102.0133 ug/L	100
97)	Hexachlorobutadiene	19.69	225	486893	102.4572 ug/L	100
98)	Naphthalene	19.90	128	2071229	103.4477 ug/L	100
99)	1,2,3-Trichlorobenzene	20.19	180	980181	99.0092 ug/L	99

(#) = qualifier out of range (m) = manual integration
11M11844.D 8260WT.M Sat May 14 18:50:58 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11844.D Vial: 9
Acq On : 13 May 2016 18:22 Operator: JDS
Sample : WG568769-09 100ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration



11M11844.D 8260WT.M Sat May 14 18:50:59 2016 Page 3

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11845.D Vial: 10
Acq On : 13 May 2016 18:54 Operator: JDS
Sample : WG568769-10 200ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50:59 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	552905	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	473364	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	274188	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	629894	104.0960 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 416.40%#		
43) 1,2-Dichloroethane-d4	10.23	65	684792	99.2511 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 397.00%#		
57) Toluene-d8	12.47	98	2068111	100.5516 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 402.20%#		
78) p-Bromofluorobenzene	15.64	95	864999	102.3411 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 409.36%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	1843265	212.7741 ug/L	99
3) Chloromethane	3.72	50	1346993	194.1434 ug/L	100
4) Vinyl Chloride	3.95	62	1199147	208.1209 ug/L	100
5) 1,3-Butadiene	3.99	54	1131910	197.5677 ug/L	97
6) Bromomethane	4.85	94	859963	217.0771 ug/L	100
7) Chloroethane	5.00	64	764060	206.3883 ug/L	100
8) Trichlorofluoromethane	5.48	101	2275404	198.1327 ug/L	100
9) Diethyl ether	6.03	59	2592	0.6103 ug/L #	21
10) Isoprene	6.04	67	1435794	202.8369 ug/L	97
11) Acrolein	6.25	56	3061	6.8231 ug/L	88
12) 1,1,2-Trichloro-1,2,2-Trif	6.25	101	1123889	201.6837 ug/L	99
13) Acetone	6.34	43	254053	193.5049 ug/L	96
14) 1,1-Dichloroethene	6.56	61	2040432	201.5858 ug/L	99
15) Tert-Butyl Alcohol	6.67	59	7729	21.2829 ug/L #	82
16) Dimethyl Sulfide	6.81	62	830179	205.3739 ug/L	99
17) Iodomethane	7.06	142	969857	206.1133 ug/L	99
18) Methyl acetate	7.07	43	749575	196.5068 ug/L	97
19) Methylene Chloride	7.31	84	1035656	197.7029 ug/L	97
20) Carbon Disulfide	7.36	76	3459748	204.1704 ug/L	99
21) Acrylonitrile	7.52	53	35662	20.2589 ug/L #	39
22) Methyl Tert Butyl Ether	7.52	73	2706665	207.1616 ug/L	100
23) trans-1,2-Dichloroethene	7.74	96	1121963	200.2808 ug/L	99
24) n-Hexane	7.82	57	1873652	200.8128 ug/L	99
25) Diisopropyl ether	8.16	45	16567	0.7116 ug/L #	87
26) Vinyl Acetate	8.31	43	1904053	202.8645 ug/L	100
27) 1,1-Dichloroethane	8.34	63	2261904	201.2314 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	9063	0.4783 ug/L	90
29) 2-Butanone	8.87	43	407543	205.3288 ug/L	99
30) Propionitrile	8.97	54	4182	7.1938 ug/L #	60
31) 2,2-Dichloropropane	9.09	77	1850091	210.3692 ug/L	100
32) cis-1,2-Dichloroethene	9.15	96	1257477	203.5843 ug/L	100
33) Chloroform	9.35	83	2160999	195.7739 ug/L	100
34) 1-Bromopropane	9.48	122	222798	201.1267 ug/L	98
35) Bromochloromethane	9.57	130	798076	217.5612 ug/L	99
36) Tetrahydrofuran	9.60	42	9110	6.4998 ug/L	97
38) 1,1,1-Trichloroethane	9.85	97	2208198	203.2875 ug/L	99
39) Cyclohexane	9.88	56	2467804	208.4625 ug/L	100
40) 1,1-Dichloropropene	10.04	75	1587009	204.6205 ug/L	99
41) Carbon Tetrachloride	10.18	117	2174730	205.0269 ug/L	99
42) Tert-Amyl-Methyl ether	10.13	73	6538	0.4965 ug/L #	83

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11845.D Vial: 10
Acq On : 13 May 2016 18:54 Operator: JDS
Sample : WG568769-10 200ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50:59 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Compound	R.T.	QIon	Response	Conc Unit	Qvalue
44) 1,2-Dichloroethane	10.34	62	1808450	201.7209 ug/L	99
45) Benzene	10.38	78	4119227	194.0096 ug/L	99
46) Trichloroethene	11.09	130	1367103	198.0963 ug/L	100
47) Methylcyclohexane	11.17	83	1789256	210.0812 ug/L	99
48) 1,2-Dichloropropane	11.29	63	1196100	203.5109 ug/L	99
49) 1,4-Dioxane	11.57	88	2656	94.1370 ug/L #	31
50) Bromodichloromethane	11.58	83	1721733	208.6734 ug/L	99
51) Dibromomethane	11.66	93	641583	209.4825 ug/L	100
52) 2-Chloroethyl Vinyl Ether	11.84	63	608231	209.9748 ug/L	99
53) 4-Methyl-2-Pentanone	11.87	58	342251	209.4831 ug/L	99
54) cis-1,3-Dichloropropene	12.17	75	1772737	214.8108 ug/L	100
55) Dimethyl Disulfide	12.42	79	1052498	211.9395 ug/L	98
58) Toluene	12.56	91	4699240	196.0993 ug/L	98
59) Ethyl Methacrylate	12.65	69	1101451	215.2567 ug/L	97
60) trans-1,3-Dichloropropene	12.73	75	1627488	211.5866 ug/L	99
61) 1,1,2-Trichloroethane	12.93	97	860113	202.4133 ug/L	99
62) 2-Hexanone	12.86	43	651287	208.6888 ug/L	99
63) 1,3-Dichloropropane	13.21	76	1412360	206.9756 ug/L	99
64) Tetrachloroethene	13.34	164	1133410	203.3874 ug/L	99
65) Dibromochloromethane	13.59	129	1409248	217.4313 ug/L	99
66) 1,2-Dibromoethane	13.82	107	889951	209.0389 ug/L	99
67) 1-Chlorohexane	13.89	91	1659272	210.0173 ug/L	97
68) Chlorobenzene	14.29	112	3416597	194.6214 ug/L	98
69) 1,1,1,2-Tetrachloroethane	14.32	131	1457008	205.1158 ug/L	100
70) Ethylbenzene	14.31	106	1814385	201.5551 ug/L	91
71) m-,p-Xylene	14.39	106	4230307	386.5510 ug/L	85
72) o-Xylene	14.92	106	2156790	202.5071 ug/L	96
73) Styrene	14.96	104	3630570	206.3240 ug/L	98
74) Bromoform	15.43	173	854308	217.4778 ug/L	100
75) Isopropylbenzene	15.31	105	5485636	194.7621 ug/L	97
77) 1,1,2,2-Tetrachloroethane	15.52	83	939708	220.0830 ug/L	98
79) 1,2,3-Trichloropropane	15.70	110	325518	202.0716 ug/L	99
80) trans-1,4-Dichloro-2-Buten	15.74	53	421012	202.2796 ug/L	97
81) n-Propylbenzene	15.79	91	6220435	195.8952 ug/L	95
82) Bromobenzene	15.92	156	1685980	202.3412 ug/L	99
83) 1,3,5-Trimethylbenzene	15.96	105	4875063	201.1866 ug/L	97
84) 2-Chlorotoluene	16.05	91	4287026	192.7532 ug/L	98
85) 4-Chlorotoluene	16.09	91	4030756	202.3214 ug/L	98
86) a-Methylstyrene	16.34	118	2724122	213.1015 ug/L	98
87) tert-Butylbenzene	16.40	134	1075661	209.5280 ug/L	96
88) 1,2,4-Trimethylbenzene	16.44	105	4921395	199.4496 ug/L	97
89) sec-Butylbenzene	16.65	105	5740919	197.3104 ug/L	97
90) p-Isopropyltoluene	16.79	119	5274037	197.3451 ug/L	95
91) 1,3-Dichlorobenzene	16.98	146	3171808	197.0945 ug/L	99
92) 1,4-Dichlorobenzene	17.10	146	3145357	192.8244 ug/L	99
93) n-Butylbenzene	17.28	91	4684496	198.9203 ug/L	97
94) 1,2-Dichlorobenzene	17.57	146	2909224	198.4498 ug/L	99
95) 1,2-Dibromo-3-Chloropropan	18.49	75	191088	196.6217 ug/L	99
96) 1,2,4-Trichlorobenzene	19.55	180	2198003	204.9021 ug/L	99
97) Hexachlorobutadiene	19.69	225	969160	205.0797 ug/L	99
98) Naphthalene	19.90	128	3994903	200.6399 ug/L	98
99) 1,2,3-Trichlorobenzene	20.19	180	1999949	203.1452 ug/L	99

(#) = qualifier out of range (m) = manual integration
11M11845.D 8260WT.M Sat May 14 18:51:00 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11845.D Vial: 10
Acq On : 13 May 2016 18:54 Operator: JDS
Sample : WG568769-10 200ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:50 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11846.D Vial: 11
Acq On : 13 May 2016 19:26 Operator: JDS
Sample : WG568769-11 300ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:51:01 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	543599	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	479542	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	277309	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	905755	152.2472 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 609.00%#		
43) 1,2-Dichloroethane-d4	10.23	65	970758	143.1066 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 572.44%#		
57) Toluene-d8	12.47	98	3002456	144.0987 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 576.40%#		
78) p-Bromofluorobenzene	15.64	95	1285822	150.4179 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 601.68%#		

Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	2762707	324.3678 ug/L	99
3) Chloromethane	3.72	50	2113477	309.8323 ug/L	100
4) Vinyl Chloride	3.95	62	1802126	318.1269 ug/L	100
5) 1,3-Butadiene	3.99	54	1562377	277.3715 ug/L	98
6) Bromomethane	4.85	94	1341865	344.5204 ug/L	100
7) Chloroethane	5.00	64	1181546	324.6236 ug/L	100
8) Trichlorofluoromethane	5.48	101	3351045	296.7904 ug/L	99
9) Diethyl ether	6.01	59	1250668	299.5049 ug/L	97
10) Isoprene	6.04	67	2121830	304.8859 ug/L	97
11) Acrolein	6.23	56	87016	155.2272 ug/L	99
12) 1,1,2-Trichloro-1,2,2-Trif	6.25	101	1679276	306.5076 ug/L	99
13) Acetone	6.34	43	352141	272.8074 ug/L	95
14) 1,1-Dichloroethene	6.56	61	3009080	302.3733 ug/L	98
15) Tert-Butyl Alcohol	6.67	59	227782	637.9680 ug/L	97
16) Dimethyl Sulfide	6.81	62	1199487	301.8151 ug/L	96
17) Iodomethane	7.06	142	1359092	295.3515 ug/L	96
18) Methyl acetate	7.07	43	1093862	303.3719 ug/L	97
19) Methylene Chloride	7.31	84	1550974	301.1438 ug/L	91
20) Carbon Disulfide	7.36	76	5025219	301.6304 ug/L	99
21) Acrylonitrile	7.49	53	292311	168.8994 ug/L	94
22) Methyl Tert Butyl Ether	7.52	73	3880019	302.0512 ug/L	100
23) trans-1,2-Dichloroethene	7.74	96	1693308	307.4459 ug/L	98
24) n-Hexane	7.82	57	2752540	300.0599 ug/L	99
25) Diisopropyl ether	8.15	45	6621718	289.2841 ug/L	99
26) Vinyl Acetate	8.31	43	2814279	303.7518 ug/L	99
27) 1,1-Dichloroethane	8.34	63	3313473	299.8312 ug/L	98
28) Ethyl-Tert-Butyl ether	8.70	59	5508427	295.6672 ug/L	99
29) 2-Butanone	8.87	43	590742	302.7235 ug/L	97
30) Propionitrile	8.99	54	177454	310.4771 ug/L	98
31) 2,2-Dichloropropane	9.09	77	2735421	316.3626 ug/L	99
32) cis-1,2-Dichloroethene	9.15	96	1856543	305.7181 ug/L	98
33) Chloroform	9.35	83	3144758	289.7739 ug/L	100
34) 1-Bromopropane	9.48	122	329211	301.9597 ug/L	98
35) Bromochloromethane	9.57	130	1179409	327.0194 ug/L	97
36) Tetrahydrofuran	9.60	42	386389	280.3985 ug/L	96
38) 1,1,1-Trichloroethane	9.85	97	3227247	302.1877 ug/L	99
39) Cyclohexane	9.88	56	3625895	311.5332 ug/L	99
40) 1,1-Dichloropropene	10.04	75	2360605	309.5743 ug/L	100
41) Carbon Tetrachloride	10.18	117	3172921	304.2542 ug/L	99
42) Tert-Amyl-Methyl ether	10.13	73	3912691	302.2388 ug/L	97

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11846.D Vial: 11
Acq On : 13 May 2016 19:26 Operator: JDS
Sample : WG568769-11 300ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:51:01 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc Unit		Qvalue
44)	1,2-Dichloroethane	10.34	62	2551762	289.5052 ug/L		98
45)	Benzene	10.38	78	5961793	285.5987 ug/L		97
46)	Trichloroethene	11.09	130	2025948	298.5901 ug/L		99
47)	Methylcyclohexane	11.17	83	2657028	317.3092 ug/L		98
48)	1,2-Dichloropropane	11.29	63	1781372	308.2809 ug/L		96
49)	1,4-Dioxane	11.56	88	21540	607.7372 ug/L		100
50)	Bromodichloromethane	11.57	83	2511276	309.5762 ug/L		100
51)	Dibromomethane	11.65	93	939605	312.0413 ug/L		99
52)	2-Chloroethyl Vinyl Ether	11.84	63	879534	308.8327 ug/L		100
53)	4-Methyl-2-Pentanone	11.87	58	505191	314.5081 ug/L		99
54)	cis-1,3-Dichloropropene	12.17	75	2604960	321.0589 ug/L		99
55)	Dimethyl Disulfide	12.42	79	1529785	313.3234 ug/L		99
58)	Toluene	12.56	91	6613670	272.4329 ug/L		95
59)	Ethyl Methacrylate	12.65	69	1618015	312.1351 ug/L		94
60)	trans-1,3-Dichloropropene	12.73	75	2369133	304.0383 ug/L		98
61)	1,1,2-Trichloroethane	12.93	97	1271039	295.2644 ug/L		98
62)	2-Hexanone	12.86	43	950079	300.5074 ug/L		97
63)	1,3-Dichloropropane	13.21	76	2062188	298.3119 ug/L		97
64)	Tetrachloroethene	13.34	164	1691190	299.5697 ug/L		98
65)	Dibromochloromethane	13.59	129	2065480	314.5750 ug/L		100
66)	1,2-Dibromoethane	13.82	107	1334668	309.4588 ug/L		99
67)	1-Chlorohexane	13.89	91	2483371	310.2757 ug/L		95
68)	Chlorobenzene	14.29	112	4911463	276.1698 ug/L		96
69)	1,1,1,2-Tetrachloroethane	14.32	131	2161701	300.4009 ug/L		100
70)	Ethylbenzene	14.31	106	2715675	297.7904 ug/L		79
71)	m-,p-Xylene	14.39	106	6004062	541.5626 ug/L		73
72)	o-Xylene	14.92	106	3202500	296.8179 ug/L		90
73)	Styrene	14.96	104	5278134	296.0902 ug/L		96
74)	Bromoform	15.43	173	1313245	330.0007 ug/L		100
75)	Isopropylbenzene	15.31	105	7610374	266.7178 ug/L		92
77)	1,1,2,2-Tetrachloroethane	15.52	83	1470787	340.5868 ug/L		97
79)	1,2,3-Trichloropropane	15.70	110	494041	303.0514 ug/L		98
80)	trans-1,4-Dichloro-2-Buten	15.74	53	640812	304.0444 ug/L		94
81)	n-Propylbenzene	15.79	91	8405785	261.7374 ug/L		88
82)	Bromobenzene	15.92	156	2530711	300.3025 ug/L		98
83)	1,3,5-Trimethylbenzene	15.96	105	6851753	279.5793 ug/L		93
84)	2-Chlorotoluene	16.05	91	6055423	269.1996 ug/L		95
85)	4-Chlorotoluene	16.09	91	5712933	283.5299 ug/L		94
86)	a-Methylstyrene	16.34	118	4020347	310.9625 ug/L		96
87)	tert-Butylbenzene	16.40	134	1654009	318.5583 ug/L		89
88)	1,2,4-Trimethylbenzene	16.44	105	6879061	275.6504 ug/L		91
89)	sec-Butylbenzene	16.65	105	7999374	271.8373 ug/L		92
90)	p-Isopropyltoluene	16.79	119	7406083	274.0036 ug/L		90
91)	1,3-Dichlorobenzene	16.98	146	4679194	287.4903 ug/L		97
92)	1,4-Dichlorobenzene	17.10	146	4642399	281.3967 ug/L		96
93)	n-Butylbenzene	17.28	91	6723928	282.3084 ug/L		93
94)	1,2-Dichlorobenzene	17.57	146	4360022	294.0672 ug/L		96
95)	1,2-Dibromo-3-Chloropropan	18.49	75	302346	307.2381 ug/L		100
96)	1,2,4-Trichlorobenzene	19.55	180	3482946	321.0328 ug/L		97
97)	Hexachlorobutadiene	19.69	225	1629665	340.9652 ug/L		98
98)	Naphthalene	19.90	128	5970381	296.4815 ug/L	#	96
99)	1,2,3-Trichlorobenzene	20.19	180	3137470	315.1025 ug/L		98

(#) = qualifier out of range (m) = manual integration

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11846.D Vial: 11
Acq On : 13 May 2016 19:26 Operator: JDS
Sample : WG568769-11 300ug/L ICAL STD 8260 Inst : hpms11
Misc : 1,1 STD76127 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:51 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration





Method Name: C:\MSDCHEM\1\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:45:57 2016



Method Name: C:\MSDCHEM\1\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:45:57 2016



Method Name: C:\MSDCHEM\1\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:45:57 2016



Method Name: C:\MSDCHEM\1\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:45:57 2016



Method Name: C:\MSDCHEM\1\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:45:57 2016



Method Name: C:\MSDCHEM\1\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:45:57 2016



Method Name: C:\MSDCHEM\1\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:45:57 2016



Method Name: C:\MSDCHEM\1\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:45:57 2016



Method Name: C:\MSDCHEM\1\METHODS\8260WT.M
Calibration Table Last Updated: Sat May 14 18:45:57 2016

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11848.D Vial: 13
Acq On : 13 May 2016 20:30 Operator: JDS
Sample : WG568769-12 50ug/L ALT SRC STD 8260 Inst : hpms11
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:47:09 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	559172	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	461765	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	267551	25.00 ug/L	0.00
System Monitoring Compounds					
37) Dibromofluoromethane	9.63	111	155853	25.4675 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 101.88%		
43) 1,2-Dichloroethane-d4	10.23	65	160000	22.9299 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 91.72%		
57) Toluene-d8	12.47	98	531547	26.4930 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 105.96%		
78) p-Bromofluorobenzene	15.64	95	203747	24.7040 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 98.80%		
Target Compounds					Qvalue
2) Dichlorodifluoromethane	3.27	85	495777	56.5878 ug/L	98
3) Chloromethane	3.72	50	386673	55.1069 ug/L	100
4) Vinyl Chloride	3.95	62	326996	56.1165 ug/L	100
5) 1,3-Butadiene	3.99	54	212443	36.6650 ug/L	98
6) Bromomethane	4.84	94	202280	50.4885 ug/L	99
7) Chloroethane	5.00	64	220373	58.8601 ug/L	99
8) Trichlorofluoromethane	5.48	101	552017	47.5286 ug/L	99
9) Diethyl ether	6.01	59	453011	105.4639 ug/L	97
10) Isoprene	6.04	67	388108	54.2141 ug/L	93
11) Acrolein	6.23	56	56029	97.7296 ug/L	98
12) 1,1,2-Trichloro-1,2,2-Trif	6.25	101	292836	51.9609 ug/L	99
13) Acetone	6.35	43	65954	49.6723 ug/L	97
14) 1,1-Dichloroethene	6.55	61	486567	47.5319 ug/L	96
15) Tert-Butyl Alcohol	6.67	59	79387	216.1534 ug/L	96
16) Dimethyl Sulfide	6.81	62	290204	70.9875 ug/L	94
17) Iodomethane	7.06	142	176766	38.4128 ug/L	92
18) Methyl acetate	7.07	43	200985	48.7416 ug/L	98
19) Methylene Chloride	7.31	84	275442	51.9916 ug/L	91
20) Carbon Disulfide	7.35	76	765803	44.6859 ug/L	99
21) Acrylonitrile	7.49	53	93674	52.6181 ug/L	100
22) Methyl Tert Butyl Ether	7.52	73	687029	51.9942 ug/L	100
23) trans-1,2-Dichloroethene	7.74	96	295337	52.1295 ug/L	97
24) n-Hexane	7.82	57	420811	44.5959 ug/L	99
25) Diisopropyl ether	8.15	45	2449232	104.0201 ug/L	98
26) Vinyl Acetate	8.31	43	395913	43.6417 ug/L	99
27) 1,1-Dichloroethane	8.34	63	551372	48.5033 ug/L	99
28) Ethyl-Tert-Butyl ether	8.70	59	1884917	98.3560 ug/L	99
29) 2-Butanone	8.87	43	101045	50.3380 ug/L	98
30) Propionitrile	8.99	54	62190	105.7785 ug/L	98
31) 2,2-Dichloropropane	9.09	77	405332	45.5727 ug/L	100
32) cis-1,2-Dichloroethene	9.15	96	332086	53.1618 ug/L	94
33) Chloroform	9.35	83	531122	47.5773 ug/L	100
34) 1-Bromopropane	9.48	122	76460	68.6656 ug/L	96
35) Bromochloromethane	9.57	130	199933	53.8923 ug/L	96
36) Tetrahydrofuran	9.60	42	131835	93.0068 ug/L	96
38) 1,1,1-Trichloroethane	9.85	97	527857	48.0501 ug/L	99
39) Cyclohexane	9.88	56	601180	50.2142 ug/L	99
40) 1,1-Dichloropropene	10.04	75	393057	50.1107 ug/L	100
41) Carbon Tetrachloride	10.18	117	507914	47.3479 ug/L	100
42) Tert-Amyl-Methyl ether	10.13	73	1406419	105.6143 ug/L	97

(#) = qualifier out of range (m) = manual integration
11M11848.D 8260WT.M Sat May 14 18:47:10 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11848.D Vial: 13
Acq On : 13 May 2016 20:30 Operator: JDS
Sample : WG568769-12 50ug/L ALT SRC STD 8260 Inst : hpms11
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:47:09 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

	Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
44)	1,2-Dichloroethane	10.34	62	434117	47.8802	ug/L	99
45)	Benzene	10.38	78	1098993	51.1808	ug/L	100
46)	Trichloroethene	11.08	130	365193	52.3242	ug/L	99
47)	Methylcyclohexane	11.17	83	450581	52.3110	ug/L	97
48)	1,2-Dichloropropane	11.29	63	313901	52.8102	ug/L	94
49)	1,4-Dioxane	11.56	88	6366	191.2043	ug/L	95
50)	Bromodichloromethane	11.57	83	413329	49.5339	ug/L	99
51)	Dibromomethane	11.65	93	148822	48.0471	ug/L	97
52)	2-Chloroethyl Vinyl Ether	11.84	63	152083	51.9140	ug/L	99
53)	4-Methyl-2-Pentanone	11.87	58	86849	52.5623	ug/L	99
54)	cis-1,3-Dichloropropene	12.17	75	480729	57.5993	ug/L	99
55)	Dimethyl Disulfide	12.42	79	261188	52.0055	ug/L	96
58)	Toluene	12.56	91	1256481	53.7500	ug/L	100
59)	Ethyl Methacrylate	12.65	69	284944	57.0854	ug/L	90
60)	trans-1,3-Dichloropropene	12.73	75	391013	52.1118	ug/L	97
61)	1,1,2-Trichloroethane	12.93	97	218717	52.7643	ug/L	99
62)	2-Hexanone	12.86	43	158873	52.1857	ug/L	98
63)	1,3-Dichloropropane	13.21	76	374611	56.2767	ug/L	95
64)	Tetrachloroethene	13.34	164	283714	52.1906	ug/L	97
65)	Dibromochloromethane	13.59	129	332092	52.5252	ug/L	99
66)	1,2-Dibromoethane	13.82	107	218604	52.6373	ug/L	99
67)	1-Chlorohexane	13.89	91	420217	54.5237	ug/L	92
68)	Chlorobenzene	14.29	112	912650	53.2936	ug/L	100
69)	1,1,1,2-Tetrachloroethane	14.32	131	353663	51.0389	ug/L	98
70)	Ethylbenzene	14.31	106	460093	52.3943	ug/L	99
71)	m-,p-Xylene	14.39	106	1144896	107.2445	ug/L	98
72)	o-Xylene	14.92	106	568307	54.7003	ug/L	100
73)	Styrene	14.95	104	943635	54.9735	ug/L	98
74)	Bromoform	15.43	173	190959	49.8328	ug/L	99
75)	Isopropylbenzene	15.31	105	1490416	54.2449	ug/L	100
77)	1,1,2,2-Tetrachloroethane	15.52	83	228913	54.9422	ug/L	100
79)	1,2,3-Trichloropropane	15.70	110	78385	50.1407	ug/L	92
80)	trans-1,4-Dichloro-2-Buten	15.74	53	73730	36.9129	ug/L	62
81)	n-Propylbenzene	15.79	91	1753049	56.5769	ug/L	99
82)	Bromobenzene	15.92	156	419932	51.6479	ug/L	99
83)	1,3,5-Trimethylbenzene	15.95	105	1303495	55.1277	ug/L	99
84)	2-Chlorotoluene	16.05	91	1145029	52.7599	ug/L	99
85)	4-Chlorotoluene	16.09	91	1077116	55.4064	ug/L	99
86)	a-Methylstyrene	16.34	118	741671	59.4584	ug/L	99
87)	tert-Butylbenzene	16.40	134	283244	56.5417	ug/L	95
88)	1,2,4-Trimethylbenzene	16.44	105	1296091	53.8297	ug/L	99
89)	sec-Butylbenzene	16.65	105	1586978	55.8961	ug/L	100
90)	p-Isopropyltoluene	16.79	119	1426648	54.7069	ug/L	99
91)	1,3-Dichlorobenzene	16.98	146	827515	52.6969	ug/L	98
92)	1,4-Dichlorobenzene	17.10	146	844174	53.0354	ug/L	98
93)	n-Butylbenzene	17.28	91	1238500	53.8957	ug/L	99
94)	1,2-Dichlorobenzene	17.57	146	766416	53.5772	ug/L	98
95)	1,2-Dibromo-3-Chloropropan	18.49	75	45158	48.1046	ug/L	97
96)	1,2,4-Trichlorobenzene	19.55	180	580278	55.4365	ug/L	99
97)	Hexachlorobutadiene	19.69	225	257859	55.9180	ug/L	98
98)	Naphthalene	19.90	128	993390	51.1296	ug/L	100
99)	1,2,3-Trichlorobenzene	20.19	180	519374	54.0642	ug/L	99

(#) = qualifier out of range (m) = manual integration
11M11848.D 8260WT.M Sat May 14 18:47:10 2016 Page 2

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\051316\11M11848.D Vial: 13
Acq On : 13 May 2016 20:30 Operator: JDS
Sample : WG568769-12 50ug/L ALT SRC STD 8260 Inst : hpms11
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: May 14 18:47 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\051316\11M11848.D Vial: 13
Acq On : 13 May 2016 20:30 Operator: JDS
Sample : WG568769-12 50ug/L ALT SRC STD 8260 Inst : hpms11
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: rteint.p

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	Fluorobenzene	25.0000	25.0000	0.0	107	0.00
2	T	Dichlorodifluoromethane	50.0000	56.5878	-13.2	119	0.00
3	P	Chloromethane	50.0000	55.1069	-10.2	134	0.00
4	C	Vinyl Chloride	50.0000	56.1165	-12.2	125	-0.01
5	T	1,3-Butadiene	50.0000	36.6650	26.7#	76	-0.01
6	T	Bromomethane	50.0000	50.4885	-1.0	117	-0.01
7	T	Chloroethane	50.0000	58.8601	-17.7	131	0.00
8	T	Trichlorofluoromethane	50.0000	47.5286	4.9	106	0.00
9	T	Diethyl ether	100.0000	105.4639	-5.5	116	0.00
10	T	Isoprene	50.0000	54.2141	-8.4	117	0.00
11	T	Acrolein	50.0000	97.7296	-95.5#	222	0.00
12	T	1,1,2-Trichloro-1,2,2-Trifl	50.0000	51.9609	-3.9	115	0.00
13	T	Acetone	50.0000	49.6723	0.7	110	0.01
14	C	1,1-Dichloroethene	50.0000	47.5320	4.9	106	-0.01
15	T	Tert-Butyl Alcohol	200.0000	216.1534	-8.1	114	0.01
16	T	Dimethyl Sulfide	50.0000	70.9875	-42.0#	150	0.00
17	T	Iodomethane	50.0000	38.4128	23.2	82	0.00
18	T	Methyl acetate	50.0000	48.7416	2.5	108	0.01
19	T	Methylene Chloride	50.0000	51.9916	-4.0	118	0.00
20	T	Carbon Disulfide	50.0000	44.6859	10.6	94	-0.01
21	T	Acrylonitrile	50.0000	52.6181	-5.2	110	0.00
22	T	Methyl Tert Butyl Ether	50.0000	51.9942	-4.0	113	0.00
23	T	trans-1,2-Dichloroethene	50.0000	52.1295	-4.3	117	0.00
24	T	n-Hexane	50.0000	44.5959	10.8	94	0.00
25	T	Diisopropyl ether	100.0000	104.0201	-4.0	112	0.00
26	T	Vinyl Acetate	50.0000	43.6417	12.7	96	0.00
27	P	1,1-Dichloroethane	50.0000	48.5033	3.0	107	0.00
28	T	Ethyl-Tert-Butyl ether	100.0000	98.3560	1.6	106	0.00
29	T	2-Butanone	50.0000	50.3381	-0.7	111	0.00
30	T	Propionitrile	100.0000	105.7786	-5.8	110	0.01
31	T	2,2-Dichloropropane	50.0000	45.5727	8.9	101	0.00
32	T	cis-1,2-Dichloroethene	50.0000	53.1618	-6.3	118	0.00
33	C	Chloroform	50.0000	47.5773	4.8	108	0.00
34	T	1-Bromopropane	50.0000	68.6656	-37.3#	152	0.00
35	T	Bromochloromethane	50.0000	53.8923	-7.8	111	0.00
36	T	Tetrahydrofuran	100.0000	93.0068	7.0	103	0.00
37	S	Dibromofluoromethane	25.0000	25.4675	-1.9	111	0.00
38	T	1,1,1-Trichloroethane	50.0000	48.0501	3.9	106	0.00
39	T	Cyclohexane	50.0000	50.2142	-0.4	105	0.00
40	T	1,1-Dichloropropene	50.0000	50.1107	-0.2	111	0.00
41	T	Carbon Tetrachloride	50.0000	47.3479	5.3	103	0.00
42	T	Tert-Amyl-Methyl ether	100.0000	105.6143	-5.6	115	0.00
43	S	1,2-Dichloroethane-d4	25.0000	22.9299	8.3	101	0.00
44	T	1,2-Dichloroethane	50.0000	47.8802	4.2	104	0.00
45	T	Benzene	50.0000	51.1809	-2.4	116	0.00
46	T	Trichloroethene	50.0000	52.3242	-4.6	119	-0.01
47	T	Methylcyclohexane	50.0000	52.3110	-4.6	108	0.00
48	C	1,2-Dichloropropane	50.0000	52.8102	-5.6	117	0.00
49	T	1,4-Dioxane	200.0000	191.2043	4.4	109	0.01
50	T	Bromodichloromethane	50.0000	49.5339	0.9	107	0.00
51	T	Dibromomethane	50.0000	48.0471	3.9	104	0.00
52	T	2-Chloroethyl Vinyl Ether	50.0000	51.9140	-3.8	114	0.00
53	T	4-Methyl-2-Pentanone	50.0000	52.5623	-5.1	115	0.00
54	T	cis-1,3-Dichloropropene	50.0000	57.5993	-15.2	122	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\051316\11M11848.D Vial: 13
Acq On : 13 May 2016 20:30 Operator: JDS
Sample : WG568769-12 50ug/L ALT SRC STD 8260 Inst : hpms11
Misc : 1,1 STD76109 Multiplr: 1.00
MS Integration Params: rteint.p

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
55	T	Dimethyl Disulfide	50.0000	52.0055	-4.0	110	0.00
56	I	Chlorobenzene-d5	25.0000	25.0000	0.0	105	0.00
57	S	Toluene-d8	25.0000	26.4930	-6.0	115	0.00
58	C	Toluene	50.0000	53.7500	-7.5	115	0.00
59	T	Ethyl Methacrylate	50.0000	57.0854	-14.2	114	0.00
60	T	trans-1,3-Dichloropropene	50.0000	52.1118	-4.2	107	0.00
61	T	1,1,2-Trichloroethane	50.0000	52.7643	-5.5	113	0.00
62	T	2-Hexanone	50.0000	52.1857	-4.4	110	0.00
63	T	1,3-Dichloropropane	50.0000	56.2767	-12.6	118	0.00
64	T	Tetrachloroethene	50.0000	52.1906	-4.4	113	0.00
65	T	Dibromochloromethane	50.0000	52.5251	-5.1	107	0.00
66	T	1,2-Dibromoethane	50.0000	52.6373	-5.3	111	0.00
67	T	1-Chlorohexane	50.0000	54.5237	-9.0	112	0.00
68	P	Chlorobenzene	50.0000	53.2936	-6.6	114	0.00
69	T	1,1,1,2-Tetrachloroethane	50.0000	51.0388	-2.1	109	0.00
70	C	Ethylbenzene	50.0000	52.3943	-4.8	113	0.00
71	T	m-,p-Xylene	100.0000	107.2445	-7.2	115	0.00
72	T	o-Xylene	50.0000	54.7003	-9.4	115	0.00
73	T	Styrene	50.0000	54.9735	-9.9	113	0.00
74	P	Bromoform	50.0000	49.8327	0.3	104	0.00
75	T	Isopropylbenzene	50.0000	54.2449	-8.5	114	0.00
76	I	1,4-Dichlorobenzene-d4	25.0000	25.0000	0.0	101	0.00
77	P	1,1,2,2-Tetrachloroethane	50.0000	54.9422	-9.9	110	0.00
78	S	p-Bromofluorobenzene	25.0000	24.7040	1.2	108	0.00
79	T	1,2,3-Trichloropropane	50.0000	50.1407	-0.3	108	0.00
80	T	trans-1,4-Dichloro-2-Butene	50.0000	36.9129	26.2#	78	0.00
81	T	n-Propylbenzene	50.0000	56.5769	-13.2	115	0.00
82	T	Bromobenzene	50.0000	51.6479	-3.3	111	0.00
83	T	1,3,5-Trimethylbenzene	50.0000	55.1277	-10.3	114	-0.01
84	T	2-Chlorotoluene	50.0000	52.7599	-5.5	112	0.00
85	T	4-Chlorotoluene	50.0000	55.4064	-10.8	116	0.00
86	T	a-Methylstyrene	50.0000	59.4584	-18.9	114	0.00
87	T	tert-Butylbenzene	50.0000	56.5417	-13.1	117	0.00
88	T	1,2,4-Trimethylbenzene	50.0000	53.8297	-7.7	110	0.00
89	T	sec-Butylbenzene	50.0000	55.8961	-11.8	114	0.00
90	T	p-Isopropyltoluene	50.0000	54.7069	-9.4	111	0.00
91	T	1,3-Dichlorobenzene	50.0000	52.6969	-5.4	112	0.00
92	T	1,4-Dichlorobenzene	50.0000	53.0354	-6.1	114	0.00
93	T	n-Butylbenzene	50.0000	53.8957	-7.8	110	0.00
94	T	1,2-Dichlorobenzene	50.0000	53.5772	-7.2	112	0.00
95	T	1,2-Dibromo-3-Chloropropane	50.0000	48.1046	3.8	105	0.00
96	T	1,2,4-Trichlorobenzene	50.0000	55.4365	-10.9	113	0.00
97	T	Hexachlorobutadiene	50.0000	55.9180	-11.8	118	0.00
98	T	Naphthalene	50.0000	51.1296	-2.3	103	0.00
99	T	1,2,3-Trichlorobenzene	50.0000	54.0642	-8.1	114	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0

2.1.1.5 Raw QC Data

BFB

```
Data File : C:\MSDCHEM\1\DATA\051316\11M11836.D          Vial: 2
Acq On    : 13 May 2016  14:15                        Operator: JDS
Sample    : WG568769-01 50ng BFB STD 8260             Inst    : hpms11
Misc      : 1,1 STD76034                              Multiplr: 1.00
MS Integration Params: rteint.p
Method    : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title     : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
```



AutoFind: Scans 174, 175, 176; Background Corrected with Scan 169

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.2	11638	PASS
75	95	30	60	50.2	24088	PASS
95	95	100	100	100.0	48016	PASS
96	95	5	9	6.3	3008	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.2	41402	PASS
175	174	5	9	8.3	3421	PASS
176	174	95	101	97.3	40288	PASS
177	176	5	9	6.9	2789	PASS

11M11836.D 8260WT.M Tue May 17 14:18:03 2016

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12803.D Vial: 6
Acq On : 22 Jun 2016 17:47 Operator: JDS
Sample : WG573776-01 BLANK STD 8260 Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:57:41 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	634960	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	558566	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	320378	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	175648	25.2763 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 101.12%		
43) 1,2-Dichloroethane-d4	10.23	65	185042	23.3535 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 93.40%		
57) Toluene-d8	12.47	98	606105	24.9737 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 99.88%		
78) p-Bromofluorobenzene	15.64	95	228605	23.1476 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 92.60%		

Target Compounds					Qvalue
3) Chloromethane	3.71	50	1290	0.1619 ug/L	100
6) Bromomethane	4.85	94	1156	0.2541 ug/L	88
13) Acetone	6.36	43	196	0.1300 ug/L #	49
16) Dimethyl Sulfide	6.81	62	689	0.1484 ug/L #	1
36) Tetrahydrofuran	9.60	42	1673	1.0394 ug/L #	63

(#) = qualifier out of range (m) = manual integration
11M12803.D 8260WT.M Fri Jun 24 14:57:42 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12803.D Vial: 6
Acq On : 22 Jun 2016 17:47 Operator: JDS
Sample : WG573776-01 BLANK STD 8260 Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:57 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration





#3
Chloromethane
Concen: 0.16 ug/L
RT: 3.71 min Scan# 110
Delta R.T. -0.01 min
Lab File: 11M12803.D
Acq: 22 Jun 2016 17:47

Tgt Ion: 50 Resp: 1290

Ion	Ratio	Lower	Upper
50	100		
52	33.6	20.0	46.6





#6
Bromomethane
Concen: 0.25 ug/L
RT: 4.85 min Scan# 220
Delta R.T. 0.00 min
Lab File: 11M12803.D
Acq: 22 Jun 2016 17:47

Tgt Ion: 94 Resp: 1156

Ion	Ratio	Lower	Upper
94	100		
96	82.4	56.5	131.9





#13
Acetone
Concen: 0.13 ug/L
RT: 6.36 min Scan# 366
Delta R.T. 0.02 min
Lab File: 11M12803.D
Acq: 22 Jun 2016 17:47

Tgt Ion: 43 Resp: 196

Ion	Ratio	Lower	Upper
43	100		
58	0.0	15.8	36.8#





#16
Dimethyl Sulfide
Concen: 0.15 ug/L
RT: 6.81 min Scan# 410
Delta R.T. 0.00 min
Lab File: 11M12803.D
Acq: 22 Jun 2016 17:47

Tgt Ion: 62 Resp: 689

Ion	Ratio	Lower	Upper
62	100		
47	0.0	73.4	171.2#





#36
Tetrahydrofuran
Concen: 1.04 ug/L
RT: 9.60 min Scan# 679
Delta R.T. 0.00 min
Lab File: 11M12803.D
Acq: 22 Jun 2016 17:47

Tgt Ion: 42 Resp: 1673

Ion	Ratio	Lower	Upper
42	100		
71	0.0	16.4	38.2#
72	16.3	16.6	38.8#



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062316\11M12831.D Vial: 5
Acq On : 23 Jun 2016 17:03 Operator: JDS
Sample : WG573953-01 BLANK STD 8260 Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 15:00:03 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	499965	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	439921	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	257668	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	142460	26.0358 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 104.16%		
43) 1,2-Dichloroethane-d4	10.23	65	155998	25.0038 ug/L	0.00
Spiked Amount 25.000	Range 80 - 120		Recovery = 100.00%		
57) Toluene-d8	12.47	98	478018	25.0080 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 100.04%		
78) p-Bromofluorobenzene	15.64	95	186251	23.4488 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 93.80%		

Target Compounds					Qvalue
13) Acetone	6.36	43	189	0.1592 ug/L #	49
36) Tetrahydrofuran	9.60	42	1570	1.2388 ug/L #	47

(#) = qualifier out of range (m) = manual integration
11M12831.D 8260WT.M Fri Jun 24 15:00:04 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062316\11M12831.D Vial: 5
Acq On : 23 Jun 2016 17:03 Operator: JDS
Sample : WG573953-01 BLANK STD 8260 Inst : hpms11
Misc : 1,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 15:00 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration





#13
Acetone
Concen: 0.16 ug/L
RT: 6.36 min Scan# 366
Delta R.T. 0.02 min
Lab File: 11M12831.D
Acq: 23 Jun 2016 17:03

Tgt Ion: 43 Resp: 189

Ion	Ratio	Lower	Upper
43	100		
58	0.0	15.8	36.8#





#36
Tetrahydrofuran
Concen: 1.24 ug/L
RT: 9.60 min Scan# 679
Delta R.T. 0.00 min
Lab File: 11M12831.D
Acq: 23 Jun 2016 17:03

Tgt Ion: 42 Resp: 1570

Ion	Ratio	Lower	Upper
42	100		
71	0.0	16.4	38.2#
72	0.0	16.6	38.8#



Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12821.D Vial: 24
Acq On : 23 Jun 2016 3:22 Operator: JDS
Sample : WG573776-06 BLANK2 STD 624 Inst : hpms11
Misc : 2,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58:49 2016 Quant Results File: 8260WT.RES

Quant Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration
DataAcq Meth : 8260WT

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene	10.61	96	472579	25.00 ug/L	0.00
56) Chlorobenzene-d5	14.25	117	423378	25.00 ug/L	0.00
76) 1,4-Dichlorobenzene-d4	17.06	152	244523	25.00 ug/L	0.00

System Monitoring Compounds

37) Dibromofluoromethane	9.63	111	137765	26.6368 ug/L	0.00
Spiked Amount 25.000	Range 86 - 118		Recovery = 106.56%		
43) 1,2-Dichloroethane-d4	10.24	65	154837	26.2559 ug/L	0.01
Spiked Amount 25.000	Range 80 - 120		Recovery = 105.04%		
57) Toluene-d8	12.47	98	463037	25.1708 ug/L	0.00
Spiked Amount 25.000	Range 88 - 110		Recovery = 100.68%		
78) p-Bromofluorobenzene	15.64	95	180039	23.8852 ug/L	0.00
Spiked Amount 25.000	Range 86 - 115		Recovery = 95.56%		

Target Compounds					Qvalue
36) Tetrahydrofuran	9.60	42	1098	0.9166 ug/L #	47

(#) = qualifier out of range (m) = manual integration
11M12821.D 8260WT.M Fri Jun 24 14:58:50 2016 Page 1

Quantitation Report (Not Reviewed)

Data File : C:\MSDCHEM\1\DATA\062216\11M12821.D Vial: 24
Acq On : 23 Jun 2016 3:22 Operator: JDS
Sample : WG573776-06 BLANK2 STD 624 Inst : hpms11
Misc : 2,1 Multiplr: 1.00
MS Integration Params: rteint.p
Quant Time: Jun 24 14:58 2016 Quant Results File: 8260WT.RES

Method : C:\MSDCHEM\1\METHODS\8260WT.M (RTE Integrator)
Title : 8260B/624 (SOP: OVL MSV01) Water 051316 HPMS11
Last Update : Sat May 14 18:45:57 2016
Response via : Initial Calibration



11M12821.D 8260WT.M Fri Jun 24 14:58:50 2016 Page 2



#36
Tetrahydrofuran
Concen: 0.92 ug/L
RT: 9.60 min Scan# 679
Delta R.T. 0.00 min
Lab File: 11M12821.D
Acq: 23 Jun 2016 3:22

Tgt Ion: 42 Resp: 1098

Ion	Ratio	Lower	Upper
42	100		
71	0.0	16.4	38.2#
72	0.0	16.6	38.8#



2.2 Semivolatiles Data

2.2.1 Semivolatiles GC/MS Data (Dioxane)

2.2.1.1 Summary Data

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	129-061616	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 18:08
Collect Date:	06/16/2016 09:05	**Dilution:**	1	**File ID:**	15M18521
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	6.52		2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	115	20	129	

Certificate of Analysis

Sample #:	L16061026-05	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	120-061616	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/27/2016 15:30
Collect Date:	06/16/2016 09:25	**Dilution:**	40	**File ID:**	15M18600
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	220		80.0	40.0	20.0

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	0.000	20	129	*

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	HPMS15
Client ID:	123-061616	**Prep Method:**	3520C	**Prep Date:**	06/21/2016 17:01
Matrix:	Water	**Analytical Method:**	8270D	**Cal Date:**	06/21/2016 11:57
Workgroup #:	WG573853	**Analyst:**	MES	**Run Date:**	06/23/2016 18:54
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	15M18523
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
1,4-Dioxane	123-91-1	0.587	J	2.00	1.00	0.500

Surrogate	Recovery	Lower Limit	Upper Limit	Q
1,4-Dioxane-d8	50.8	20	129	

J	Estimated value ; the analyte concentration was less than the LOQ.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.2.1.2 QC Summary Data

Example 8270 Calculations

1.0 Calculating the Response Factor (RF) from the initial calibration (ICAL) data:

RF = [(Ax) (Cis)] / [(Ais) (Cx)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured:	1261197
Cis = Concentration of the specific internal standard (ug/mL)	40
Ais = Area of the characteristic ion of the specific internal standard	608044
Cx = Concentration of the compound in the standard being measured (ug/mL)	50
RF = Calculated Response Factor	**1.65935**

2.0 Calculating the concentration (C) of a compound in water using the data from the prep log and quantitation report: *

Cx = [(Ax) (Cis) (Vf) (D)] / [(Ais) (RF) (Vi)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured	367250
Cis = Concentration of the specific internal standard (ug/mL)	40
Vf = Final volume of sample extract from prep log (mL)	1
D = Dilution factor for sample as a multiplier (10x = 10)	1
Ais = Area of the characteristic ion of the specific internal standard	511641
RF = Average RF from the ICAL	1.65935
Vi = Initial volume of sample extracted from prep log (mL)	1021
Cx = Concentration of the compound in the sample being measured (ug/mL)	0.016947
Cx = Concentration of the compound in the sample being measured (ug/L)	**16.947**

3.0 Calculating the concentration (C) of a compound in soil using the data from the prep log and quantitation report: *

Cx = [(Ax) (Cis) (Vf) (D)] / [(Ais) (RF) (Wi)]

where:	Example
Ax = Area of the characteristic ion for the compound being measured	367250
Cis = Concentration of the specific internal standard (ug/mL)	40
Vf = Final volume of sample extract from prep log (mL)	1
D = Dilution factor for sample as a multiplier (10x = 10)	1
Ais = Area of the characteristic ion of the specific internal standard	511641
RF = Average RF from the ICAL	1.65935
Wi = Initial weight of sample extracted (g) from prep log	30
Cx = Concentration of the compound in the sample being measured (ug/g)	0.576763
Cx = Concentration of the compound in the sample being measured (ug/kg)	**576.7627**

Dry weight correction:

Percent solids (PCT_S)	50
Cd = (Cx) (100)/PCT_S	**1153.525** ug/kg

* Concentrations appearing on the instrument quantitation reports are on-column results and do not take into account initial volume, final volume, and the dilution factor.

4.0 Concentration from Linear Regression

Step 1: Retrieve Curve Data From Plot, y = mx + b

y = response ratio = response of analyte / response of IS = Ax/Ais

x = amount ratio = concentration analyte/concentration internal standard = Cx / Cis

m = slope from curve plot

b = intercept from curve plot

Step 2: Calculate y from Quantitation Report

y = 16790/784838 = 0.02139

Step 3: Solve for x

x = (y - b)/m = [(0.02139 -(- 0.0435)]/ 0.0783 = 0.829

Step 4: Solve for analyte concentration Cx

Cx = Cis (x) = (25.0)(0.829) = 20.72 ug/L

Example Spreadsheet Calculation:

Slope from curve, m:	0.0783	
Intercept from curve, b:	-0.0435	
Area of analyte, Ax:	16790	
Area of Internal Standard , Ais:	784484	
Concentration of IS, Cis	25.00	ug/L
Response Ratio (y) :	0.021403	
Amount Ratio:	0.828897	
Concentration (Cx):	20.72241	ug/L

5.0 Concentration from Quadratic Regression

Step 1 - Retrieve Curve Data from Plot, y = Ax^2 + Bx + C
Where:
Ax^2 + Bx + (C - y) = 0
A, B, C = constants from the ICAL quadratic regression
y = Response ratio = Area of analyte/Area of internal standard (IS)
x = Amount ratio = Concentration of analyte/concentration of IS

Step 2: Calculate y from Quantitation Report

y = Ax/Ais

Step 3: Solve for x using the quadratic formula
Ax^2 + Bx + C - y = 0

$$x = \frac{b \pm \sqrt{\left(b^2 - 4a(c - y)\right)}}{2a}$$ (Two possible solutions)

Step 4: Solve for analyte concentration Cx

Cx = (Cis)(Amount ratio)

Example Spreadsheet Calculation:

Value of A from plot:	**0.0259**	
Value of B from plot:	**0.0596**	
Value of C from plot:	**-0.0165**	
Area of analyte from quantitation report:	**203233**	
Area of IS from quantiation report:	**1425653**	
Response ratio, y:	**0.142554**	
C - y:	**-0.15905**	
Root 1 - Computed amount ratio , X1:	**-3.88278**	
Root 2 - Computed amount ratio , X2:	**1.581623**	use this solution
Concentration of IS, Cis:	**40.00**	
Concentration of analyte, Cx:	**63.26**	ug/L

Microbac Laboratories Inc.
Continuous Sample Extract Log

Workgroup: WG573299
Analyst: CPD
Spike Analyst: CPD
Method: 3520C
Run Date: 06/21/2016 17:01
SOP: EXB01 Revison 20
Spike Witness: JLD
Surr Solution: STD76628

TIME ON: 17:45 OFF: 12:10 ON: 16:00 OFF: 11:05

CH2CL2 Lot #: COA18982
sodium sulfate Lot #: COA18814
10N NaOH Lot #: RGT37076
1:1 H2SO4 Lot #: RGT36755

	SAMPLE #	Type	Reference	Prod	pH	Init Amnt	Surr Amnt	Spike Amnt	Spike Sol	Final Vol	Color
1	L16060927-01	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
2	L16060927-02	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
3	L16060927-03	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
4	L16060927-04	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
5	L16060930-03	SAMP		827-DIOX	<2>12	1000 mL	.05 mL			1 mL	Colored
6	L16060978-03	SAMP		827-DIOX	<2>12	800 mL	.05 mL			1 mL	Colored
7	L16060978-04	SAMP		827-DIOX	<2>12	890 mL	.05 mL			1 mL	Colored
8	L16061026-04	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
9	L16061026-05	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
10	L16061026-07	SAMP		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
11	L16061055-04	SAMP		827-DIOX	<2>12	520 mL	.05 mL			1 mL	Colored
12	L16061055-07	SAMP		827-DIOX	<2>12	880 mL	.05 mL			1 mL	Colored
13	WG573299-01	BLANK		827-DIOXANE	<2>12	1000 mL	.05 mL			1 mL	Transparent
14	WG573299-02	LCS		827-DIOXANE	<2>12	1000 mL	.05 mL	.1 mL	STD75281	1 mL	Transparent
15	WG573299-03	LCS2		827-DIOXANE	<2>12	1000 mL	.05 mL	.1 mL	STD75281	1 mL	Transparent

Due to insufficient sample volume, this preparation batch failed to include the method prescribed MS and MSD.

Analyst: Charlie Davis

Reviewer: Justin Hesson

EXTRACT - Modified 01/23/2009
PDF ID: 4819879
Report generated: 06/24/2016 10:45

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062116
Analyst1: MES Analyst2: NA
Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:
Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG573365

Internal STD: STD76627 Surrogate STD: NA Calibration STD

CCV STD: STD74784 LCS STD: STD74214 MS/MSD STD:

Comments:

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	15M18455	BAKEOUT	1	1		06/21/16 08:46
2	15M18456	WG573501-01 5PPM DFTPP STD	1	1	STD75016	06/21/16 09:05
3	15M18457	WG573501-01 5PPM DFTPP STD	1	1	STD75016	06/21/16 09:34
4	15M18458	WG573501-02 5PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 09:52
5	15M18459	WG573501-03 10 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 10:27
6	15M18460	WG573501-04 7.5 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 10:49
7	15M18461	WG573501-05 2.5 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 11:12
8	15M18462	WG573501-06 1 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 11:34
9	15M18463	WG573501-07 0.4 PPM 1,4-DIOXANE STD	1	1	STD74784	06/21/16 11:57
10	15M18464	WG573501-08 5PPM 1,4-DIOXANE ALT SRC	1	1	STD74214	06/21/16 12:20
11	15M18465	WG572806-01 BLANK 6/15	1	1		06/21/16 12:54
12	15M18466	WG572806-02 LCS 6/15	1	1		06/21/16 13:17
13	15M18467	WG572806-03 LCSDUP 6/15	1	1		06/21/16 13:40
14	15M18468	WG573008-01 BLANK 6/17	7	1	SOIL	06/21/16 14:03
15	15M18469	WG573008-02 LCS 6/17	7	1	SOIL	06/21/16 14:26
16	15M18470	WG573008-03 LCSDUP 6/17	7	1	SOIL	06/21/16 14:48
17	15M18471	L16060730-01	1	1		06/21/16 15:11
18	15M18472	L16060730-02	1	1		06/21/16 15:34
19	15M18473	L16060854-01	1	1		06/21/16 15:57
20	15M18474	L16060730-03	1	1		06/21/16 16:19
21	15M18475	L16060756-04 827-DIOX	1	1		06/21/16 16:46
22	15M18476	L16060822-01	1	1		06/21/16 17:08
23	15M18477	L16060822-03	1	1		06/21/16 17:31
24	15M18478	L16060822-04	1	1		06/21/16 17:54
25	15M18479	L16060822-05	1	1		06/21/16 18:17
26	15M18480	L16060823-01	1	1		06/21/16 18:40
27	15M18481	L16060823-03	1	1		06/21/16 19:02
28	15M18482	L16060823-04	1	1		06/21/16 19:25
29	15M18483	L16060809-03 827-DIOX	1	1		06/21/16 19:48
30	15M18484	L16060854-01 2X	1	2		06/21/16 20:10
31	15M18485	BAKEOUT	1	1		06/21/16 20:33
32	15M18486	BAKEOUT	1	1		06/21/16 20:55
33	15M18487	BAKEOUT	1	1		06/21/16 21:18

Page: 1 Approved: 22-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062116

Analyst1: MES Analyst2: NA

Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:

Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG573365

Internal STD: STD76627 Surrogate STD: NA

CCV STD: STD74784 LCS STD: STD74214

Comments

Seq.	Rerun	Dil.	Reason	Analytes
2				
	WG573501-01 5PPM DFTPP STD rerun DFTPP, pentachlorophenol tailing >2.			
14				
	WG573008-01 BLANK 6/17 DNR-not needed.			
15				
	WG573008-02 LCS DNR-not needed.			
16				
	WG573008-03 LCSDUP DNR-not needed.			
19	X	2	Over Calibration Range	
	L16060854-01			

Page: 2 Approved: 22-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062316
Analyst1: MES Analyst2: NA
Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:
Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG573853

Internal STD: STD76627 Surrogate STD: NA Calibration STD
CCV STD: STD74784 LCS STD: MS/MSD STD:

Comments:

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	15M18504	BAKEOUT	1	1		06/23/16 10:23
2	15M18505	WG573879-01 5PPM DFTPP STD	1	1	STD75016	06/23/16 10:43
3	15M18506	WG573879-01 5PPM DFTPP STD	1	1	STD75016	06/23/16 10:59
4	15M18507	WG573879-01 5PPM DFTPP STD	1	1	STD75016	06/23/16 12:04
5	15M18508	WG573879-01 5PPM DFTPP STD	1	1	STD75016	06/23/16 13:20
6	15M18509	WG573879-01 5PPM DFTPP STD	1	1	STD75016	06/23/16 13:42
7	15M18510	WG573879-02 5PPM 1,4-DIOXANE STD	1	1	STD74784	06/23/16 13:59
8	15M18511	WG573299-01 BLANK 6/21	1	1		06/23/16 14:22
9	15M18512	WG573299-02 LCS 6/21	1	1		06/23/16 14:45
10	15M18513	WG573299-03 LCSDUP 6/21	1	1		06/23/16 15:07
11	15M18514	L16060927-01 827-DIOX	1	1		06/23/16 15:30
12	15M18515	L16060927-02 827-DIOX	1	1		06/23/16 15:53
13	15M18516	L16060927-03 827-DIOX	1	1		06/23/16 16:15
14	15M18517	L16060927-04 827-DIOX	1	1		06/23/16 16:38
15	15M18518	L16060930-03 827-DIOX	1	1		06/23/16 17:00
16	15M18519	L16060978-03 827-DIOX	1	1		06/23/16 17:23
17	15M18520	L16060978-04 827-DIOX	1	1		06/23/16 17:46
18	15M18521	L16061026-04 827-DIOX	1	1		06/23/16 18:08
19	15M18522	L16061026-05 827-DIOX	1	1		06/23/16 18:31
20	15M18523	L16061026-07 827-DIOX	1	1		06/23/16 18:54
21	15M18524	L16061055-04 827-DIOX	1	1		06/23/16 19:16
22	15M18525	L16061055-07 827-DIOX	1	1		06/23/16 19:39
23	15M18526	BAKEOUT	1	1		06/23/16 20:02
24	15M18527	BAKEOUT	1	1		06/23/16 20:24
25	15M18528	BAKEOUT	1	1		06/23/16 20:47

Comments

Seq.	Rerun	Dil.	Reason	Analytes
2				
	Rerun DFTPP, degradation >20%.			
3				
	Rerun DFTPP, degradation >20%. Changing the liner.			
4				

Page: 1 Approved: 24-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062316

Analyst1: MES Analyst2: NA

Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:

Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG573853

Internal STD: STD76627 Surrogate STD: NA

CCV STD: STD74784 LCS STD:

Comments

Seq.	Rerun	Dil.	Reason	Analytes
	Rerun DFTPP, degradation still >20%. Changing the gold seal.			
5				
	Rerun DFTPP, degradation still >20%. Clipping the column.			
12	X	10	Surrogate standard failure	
	L16060927-02 high surrogate due to SMI; reanalyzing at a dilution to confirm.			
16	X	10	Over Calibration Range	
	L16060978-03			
17	X	20	Over Calibration Range	
	L16060978-04			
19	X	20	Over Calibration Range	
	L16061026-05 high surrogate due to SMI.			
21	X	10	Over Calibration Range	
	L16061055-04 827-DIOX one high surrogate due to SMI.			
22	X	100	Over Calibration Range	
	L16061055-07 827-DIOX one low surrogate due to obvious SMI.			

Page: 2 Approved: 24-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062716
Analyst1: MES Analyst2: NA
Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:
Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG574301,WG574187,WG574186,WG573853

Internal STD: STD76627 Surrogate STD: NA Calibration STD

CCV STD: STD74784 LCS STD: MS/MSD STD:

Comments: All samples associated with extraction workgroup WG573815 are being re-extracted due to low LCSDUP recovery: L16060637-02,756-01 to-03,809-01 and -02,930-01 and -02, 1213-01 to -06.

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	15M18584	BAKEOUT	1	1		06/27/16 08:21
13	15M18585	WG574241-01 5PPM DFTPP STD	1	1	STD75016	06/27/16 08:41
14	15M18586	WG574241-01 5PPM DFTPP STD	1	1	STD75016	06/27/16 08:58
2	15M18587	WG574241-01 5PPM DFTPP STD	1	1	STD75016	06/27/16 09:22
3	15M18588	WG574241-01 5PPM DFTPP STD	1	1	STD75016	06/27/16 09:38
4	15M18589	WG574241-01 5PPM DFTPP STD	1	1	STD75016	06/27/16 10:23
5	15M18590	WG574241-01 5PPM DFTPP STD	1	1	STD75016	06/27/16 11:15
6	15M18591	WG574241-01 5PPM DFTPP STD	1	1	STD75016	06/27/16 11:32
7	15M18592	WG574241-01 5PPM DFTPP STD	1	1	STD75016	06/27/16 12:20
8	15M18593	WG574241-02 5PPM 1,4-DIOXANE STD	1	1	STD74784	06/27/16 12:37
9	15M18594	WG573975-03 LCSDUP 6/24 RR	7	1	SOIL	06/27/16 13:15
10	15M18595	WG573815-02 LCS 6/23 RR	7	1	SOIL	06/27/16 13:37
11	15M18596	WG573815-03 LCSDUP 6/23 RR	7	1	SOIL	06/27/16 13:59
12	15M18597	WG573812-01 BLANK 6/22	1	1		06/27/16 14:22
15	15M18598	WG573812-02 LCS 6/22	1	1		06/27/16 14:45
16	15M18599	WG573812-03 LCSDUP 6/22	1	1		06/27/16 15:07
17	15M18600	L16061026-05 40X	1	40		06/27/16 15:30
18	15M18601	L16061213-05 827-DIOX	7	1	SOIL	06/27/16 15:52
19	15M18602	L16061213-06 827-DIOX	7	1		06/27/16 16:15
20	15M18603	L16061233-02 827-DIOX	1	1		06/27/16 16:38
21	15M18604	L16061124-01	1	1		06/27/16 17:01
22	15M18605	L16061142-01	1	1		06/27/16 17:23
23	15M18606	L16061142-06	1	1		06/27/16 17:46
24	15M18607	L16061142-07	1	1		06/27/16 18:09
25	15M18608	L16061144-01	1	1		06/27/16 18:31
26	15M18609	L16061144-02	1	1		06/27/16 18:54
27	15M18610	L16061144-03	1	1		06/27/16 19:17
28	15M18611	L16061227-01	1	1		06/27/16 19:39
29	15M18612	L16061227-06	1	1		06/27/16 20:02
30	15M18613	L16061227-07	1	1		06/27/16 20:25
31	15M18614	L16061124-01 2X	1	2		06/27/16 20:47
32	15M18615	BAKEOUT	1	1		06/27/16 21:10
33	15M18616	BAKEOUT	1	1		06/27/16 21:32

Page: 1 Approved: 28-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: HPMS15 Dataset: 062716
Analyst1: MES Analyst2: NA
Method: 8270C/D SOP: MSS01 Rev: 27

Maintenance Log ID: Syringe Filter Lot#:
Eluent ID#:

Column 1 ID: RXI-5MS Column 2 ID: NA

Workgroups: WG574301,WG574187,WG574186,WG573853

Internal STD: STD76627 Surrogate STD: NA

CCV STD: STD74784 LCS STD:

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
34	15M18617	BAKEOUT	1	1		06/27/16 21:55

Comments

Seq.	Rerun	Dil.	Reason	Analytes
13				
	Rerun DFTPP, wrong standard was analyzed.			
14				
	Rerun DFTPP, wrong standard was analyzed.			
2				
	Rerun DFTPP, ion failures. Liner needs changed.			
3				
	Rerun DFTPP, degradation >20%.			
4				
	Rerun DFTPP, degradation >20%. Changed the gold seal.			
5				
	Rerun DFTPP, benzidine tailing >2.			
6				
	Rerun DFTPP, benzidine tailing >2. Clipped 5cm from the column.			
9				
	WG573975-03 LCSDUP 6/24 RR recovery is acceptable. Not reporting the run from 6/25/16.			
11				
	WG573815-03 LCSDUP 6/23 RR low recovery; confirms the original run.			
21	X	2	Over Calibration Range	
	L16061124-01			
28				
	L16061227-01 Needs re-extracted; surrogate recovery is low.			

Page: 2 Approved: 28-JUN-16

Microbac Laboratories Inc.

Data Checklist

Date: 21-JUN-2016
Analyst: MES
Analyst: NA
Method: 827-DIOX
Instrument: HPMS15
Curve Workgroup: NA
Runlog ID: 75850
Analytical Workgroups: L16060730, 060756, 060809, 060822, 060823, 060854

ANALYTICAL	
System Performance Check	X
DFTPP (MS)	X
Endrin/DDT breakdown (8081/MS)	X
Pentachlorophenol/benzidine tailing (MS)	X
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	X
Average RF	X
Linear regression or higher order curve	NA
Alternate source standard (ICV) % Difference	X
Continuing Calibration (CCV)	NA
% D/% Drift	NA
Minimum response factors (MS)	NA
Continuing calibration blank (CCB) (IC)	NA
Special standards	NA
Blanks	X
TCL hits	X
Surrogate recoveries	X
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	X
MS/MSD/Sample duplicates	NA
Recoveries	NA
%RPD	NA
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	X
Surrogate recoveries	X
Internal standard areas (MS)	X
Library searches (MS)	NA
Calculations & correct factors	X
Compounds above calibration range	X
Reruns	X
Manual integrations	NA
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	NA
Check for completeness	X
Primary Reviewer	MES
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	ECL

Primary Reviewer: 22-JUN-2016

Secondary Reviewer: 22-JUN-2016

CHECKLIST1 - Modified 03/05/2008

Generated: JUN-22-2016 16:00:51

Microbac Laboratories Inc.

Data Checklist

Date: 23-JUN-2016
Analyst: MES
Analyst: NA
Method: 827-DIOX
Instrument: HPMS15
Curve Workgroup: NA
Runlog ID: 75899
Analytical Workgroups: L16060927,060930,060978,-061026,061055

ANALYTICAL	
System Performance Check	X
DFTPP (MS)	X
Endrin/DDT breakdown (8081/MS)	X
Pentachlorophenol/benzidine tailing (MS)	X
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	NA
Average RF	NA
Linear regression or higher order curve	NA
Alternate source standard (ICV) % Difference	NA
Continuing Calibration (CCV)	X
% D/% Drift	X
Minimum response factors (MS)	X
Continuing calibration blank (CCB) (IC)	NA
Special standards	NA
Blanks	X
TCL hits	X
Surrogate recoveries	X
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	X
MS/MSD/Sample duplicates	NA
Recoveries	NA
%RPD	NA
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	X
Surrogate recoveries	X
Internal standard areas (MS)	X
Library searches (MS)	NA
Calculations & correct factors	X
Compounds above calibration range	X
Reruns	X
Manual integrations	X
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	NA
Check for completeness	X
Primary Reviewer	MES
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	WTD

Primary Reviewer: Secondary Reviewer:
24-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-24-2016 14:48:11

Microbac Laboratories Inc.

Data Checklist

Date: 27-JUN-2016
Analyst: MES
Analyst: NA
Method: 827-DIOX
Instrument: HPMS15
Curve Workgroup: NA
Runlog ID: 75949
Analytical Workgroups: L16061026,061213,WG573956-LCSDUP RR,061124,1142,1144,1227,123

ANALYTICAL	
System Performance Check	X
DFTPP (MS)	X
Endrin/DDT breakdown (8081/MS)	X
Pentachlorophenol/benzidine tailing (MS)	X
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	NA
Average RF	NA
Linear regression or higher order curve	NA
Alternate source standard (ICV) % Difference	NA
Continuing Calibration (CCV)	X
% D/% Drift	X
Minimum response factors (MS)	X
Continuing calibration blank (CCB) (IC)	NA
Special standards	NA
Blanks	X
TCL hits	X
Surrogate recoveries	X
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	X
MS/MSD/Sample duplicates	NA
Recoveries	NA
%RPD	NA
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	X
Surrogate recoveries	X
Internal standard areas (MS)	X
Library searches (MS)	NA
Calculations & correct factors	X
Compounds above calibration range	X
Reruns	NA
Manual integrations	NA
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	NA
Check for completeness	X
Primary Reviewer	MES
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	WTD

Primary Reviewer:
28-JUN-2016

Secondary Reviewer:
28-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-28-2016 14:33:27

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 8270D

AAB#: WG573853

Login Number: L16061026

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
129-061616	04	06/16/16					06/21/2016	5.3	7		06/23/16	2	40	
120-061616	05	06/16/16					06/21/2016	5.3	7		06/27/16	5.9	40	
123-061616	07	06/16/16					06/21/2016	5.3	7		06/23/16	2.1	40	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4822554
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.
SURROGATE STANDARDS

Login Number: L16061026
Instrument Id: HPMS15
Workgroup (AAB#): WG573853

Method: 827-DIOXANE
CAL ID: HPMS15-21-JUN-16
Matrix: Water

Sample Number	Dilution	Tag	1
L16061026-04	1.00	01	115
L16061026-05	40.0	DL01	DL
L16061026-07	1.00	01	50.8
WG573299-01	1.00	01	28.7
WG573299-02	1.00	01	37.8
WG573299-03	1.00	01	46.8

Surrogates	Surrogate Limits		
1 - 1,4-Dioxane-d8	20	-	129

Underline = Result out of surrogate limits

DL = surrogate diluted out
ND = surrogate not detected

SURROGATES - Modified 03/06/2008
PDF File ID: 4822562
Report generated: 06/29/2016 08:21

METHOD BLANK SUMMARY

Login Number: L16061026
Blank File ID: 15M18511
Prep Date: 06/21/16 17:01
Analyzed Date: 06/23/16 14:22
Analyst: MES

Work Group: WG573853
Blank Sample ID: WG573299-01
Instrument ID: HPMS15
Method: 8270D

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573299-02	15M18512	06/23/16 14:45	01
LCS2	WG573299-03	15M18513	06/23/16 15:07	01
129-061616	L16061026-04	15M18521	06/23/16 18:08	01
123-061616	L16061026-07	15M18523	06/23/16 18:54	01
120-061616	L16061026-05	15M18600	06/27/16 15:30	DL01

Report Name: BLANK_SUMMARY
PDF File ID: 4828845
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16061026 Prep Date: 06/21/16 17:01 Sample ID: WG573299-01
Instrument ID: HPMS15 Run Date: 06/23/16 14:22 Prep Method: 3520C
File ID: 15M18511 Analyst: MES Method: 8270D
Workgroup (AAB#): WG573853 Matrix: Water Units: ug/L
Contract #: Cal ID: HPMS15-21-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
1,4-Dioxane	0.500	2.00	0.500	1	U

Surrogates	% Recovery	Surrogate Limits	Qualifier
1,4-Dioxane-d8	28.7	20 - 129	PASS

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name:BLANK
PDF ID: 4822555
29-JUN-2016 08:21

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16061026 Analyst: MES Prep Method: 3520C
Instrument ID: HPMS15 Matrix: Water Method: 8270D
Workgroup (AAB#): WG573853 Units: ug/L
QC Key: DOD4 Lot #: STD75281

Sample ID: WG573299-02 LCS File ID: 15M18512 Run Date: 06/23/2016 14:45
Sample ID: WG573299-03 LCS2 File ID: 15M18513 Run Date: 06/23/2016 15:07

Analytes	LCS			LCS2			%RPD	%Rec Limits	RPD Lmt	Q
	Known	Found	% REC	Known	Found	% REC				
1,4-Dioxane	5.00	1.88	37.6	5.00	2.32	46.4	20.8	30 - 104	30	

Surogates	LCS	LCS2	Surrogate Limits	Qualifier
	% Recovery	% Recovery		
1,4-Dioxane-d8	37.8	46.8	20 - 129	PASS

* EXCEEDS %REC LIMIT

\# EXCEEDS RPD LIMIT

LCS_LCS2 - Modified 03/06/2008
PDF File ID: 4822556
Report generated: 06/29/2016 08:21

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

DFTPP

Login Number: L16061026
Instrument: HPMS15
Analyst: MES
Workgroup: WG573501

Tune ID: WG573501-01
Run Date: 06/21/2016
Run Time: 09:34
File ID: 15M18457
Cal ID: HPMS15 - 21-JUN-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51.0	198	30.0	60.0	38.8	34488	PASS
68.0	69.0	0	2.00	1.36	500	PASS
69.0	198	0	100	41.4	36833	PASS
70.0	69.0	0	2.00	0.467	172	PASS
127	198	40.0	60.0	51.9	46131	PASS
197	198	0	1.00	0	0	PASS
198	198	100	100	100	88880	PASS
199	198	5.00	9.00	6.66	5918	PASS
275	198	10.0	30.0	24.9	22090	PASS
365	198	1.00	100	2.65	2358	PASS
441	443	0.0100	100	75.7	9586	PASS
442	198	40.0	100	73.4	65261	PASS
443	442	17.0	23.0	19.4	12659	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573501-02	STD-CCV	01	06/21/2016 09:52	
WG573501-03	STD	01	06/21/2016 10:27	
WG573501-04	STD	01	06/21/2016 10:49	
WG573501-05	STD	01	06/21/2016 11:12	
WG573501-06	STD	01	06/21/2016 11:34	
WG573501-07	STD	01	06/21/2016 11:57	
WG573501-08	SSCV	01	06/21/2016 12:20	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4822559
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

DFTPP

Login Number: L16061026
Instrument: HPMS15
Analyst: MES
Workgroup: WG573879

Tune ID: WG573879-01
Run Date: 06/23/2016
Run Time: 13:42
File ID: 15M18509
Cal ID: HPMS15 - 21-JUN-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51.0	198	30.0	60.0	37.0	43111	PASS
68.0	69.0	0	2.00	1.10	506	PASS
69.0	198	0	100	39.4	45795	PASS
70.0	69.0	0	2.00	0.157	72	PASS
127	198	40.0	60.0	51.3	59752	PASS
197	198	0	1.00	0	0	PASS
198	198	100	100	100	116364	PASS
199	198	5.00	9.00	7.09	8249	PASS
275	198	10.0	30.0	24.9	28968	PASS
365	198	1.00	100	2.82	3278	PASS
441	443	0.0100	100	79.2	13093	PASS
442	198	40.0	100	74.3	86464	PASS
443	442	17.0	23.0	19.1	16541	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG573879-02	CCV	01	06/23/2016 13:59	
WG573299-01	BLANK	01	06/23/2016 14:22	
WG573299-02	LCS	01	06/23/2016 14:45	
WG573299-03	LCS2	01	06/23/2016 15:07	
L16061026-04	129-061616	01	06/23/2016 18:08	
L16061026-07	123-061616	01	06/23/2016 18:54	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4822559
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.
ORGANIC INSTRUMENT CHECK

DFTPP

Login Number: L16061026
Instrument: HPMS15
Analyst: MES
Workgroup: WG574241

Tune ID: WG574241-01
Run Date: 06/27/2016
Run Time: 12:20
File ID: 15M18592
Cal ID: HPMS15 - 21-JUN-16

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51.0	198	30.0	60.0	33.2	31989	PASS
68.0	69.0	0	2.00	1.61	580	PASS
69.0	198	0	100	37.4	35954	PASS
70.0	69.0	0	2.00	0.617	222	PASS
127	198	40.0	60.0	50.4	48459	PASS
197	198	0	1.00	0.425	409	PASS
198	198	100	100	100	96232	PASS
199	198	5.00	9.00	6.65	6399	PASS
275	198	10.0	30.0	25.2	24236	PASS
365	198	1.00	100	2.50	2410	PASS
441	443	0.0100	100	75.7	10371	PASS
442	198	40.0	100	73.9	71133	PASS
443	442	17.0	23.0	19.3	13708	PASS

This check relates to the following samples:

Lab ID	Client ID	Tag	Date Analyzed	Q
WG574241-02	CCV	01	06/27/2016 12:37	
L16061026-05	120-061616	DL01	06/27/2016 15:30	

* Sample past 12 hour tune limit

TUNE - Modified 03/06/2008
PDF File ID: 4822559
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.

INITIAL CALIBRATION SUMMARY

Login Number: L16061026
Analytical Method: 8270D
ICAL Workgroup: WG573501

Instrument ID: HPMS15
Initial Calibration Date: 21-JUN-16 11:57
Column ID: F

Analyte		AVG RF	% RSD	LINEAR (R)	QUAD (R^2)
1,4-Dioxane		0.3060	2.56		

R = Correlation coefficient; 0.995 minimum
R^2 = Coefficient of determination; 0.99 minimum

INT_CAL - Modified 03/06/2008
PDF File ID: 4822557
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16061026

Analytical Method: 8270D

Instrument ID: HPMS15

Initial Calibration Date: 21-JUN-16 11:57

Column ID: F

Analyte	WG573501-02			WG573501-03			WG573501-04		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
1,4-Dioxane	5.00	60883.0000	0.3073	10.0	138820.000	0.3164	7.50	103182.000	0.3106

INT_CAL - Modified 03/06/2008
PDF File ID: 4822557
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16061026

Analytical Method: 8270D

Instrument ID: HPMS15

Initial Calibration Date: 21-JUN-16 11:57

Column ID: F

Analyte	WG573501-05			WG573501-06			WG573501-07		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
1,4-Dioxane	2.50	26333.0000	0.2969	1.00	13250.0000	0.2966	0.400	5650.00000	0.3081

INT_CAL - Modified 03/06/2008
PDF File ID: 4822557
Report generated 06/29/2016 08:21

Calibration Table Report
Method: DIOXANE_D8.M
Title: OVD MSS01 SIM 1,4-dioxane ICAL 062116
Last Calibration: Tue Jun 21 12:51:17 2016
Curve:WG573501
Calibration Files

Compound	10 15M18459.D	7.5 15M18460.D	5 15M18458.D	2.5 15M18461.D	1 15M18462.D	0.4 15M18463.D	Avg	%RSD
1,4-Dichlorobenzene-d4	ISTD							
1,4-Dioxane	0.316	0.311	0.307	0.297	0.297	0.308	0.306	2.558
1,4-Dioxane-d8	0.426	0.418	0.417	0.402	0.395		0.411	3.044
Nitrobenzene-d5	1.202	1.173	1.128	1.002	0.927	0.906	1.056	12.163
2-Fluorobiphenyl	2.735	2.736	2.691	2.621	2.563	2.651	2.666	2.548
p-Terphenyl-d14	2.793	2.774	2.622	2.460	2.367	2.405	2.570	7.266

Wed Jun 22 12:17:33 2016

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16061026 Run Date: 06/21/2016 Sample ID: WG573501-08
Instrument ID: HPMS15 Run Time: 12:20 Method: 8270D
File ID: 15M18464 Analyst: MES QC Key: DOD4
ICal Workgroup: WG573501 Cal ID: HPMS15 - 21-JUN-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
1,4-Dioxane		5000	5050	ug/L	0.309	0.900	20	

* Exceeds %D Limit

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4822558
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/23/2016
Sample ID: WG573879-02
Instrument ID: HPMS15
Run Time: 13:59
Method: 8270D
File ID: 15M18510
Analyst: MES
QC Key: DOD4
Workgroup (AAB#): WG573853
Cal ID: HPMS15 - 21-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
1,4-Dioxane		5000	4530	ug/L	0.277	9.36	20	

* Exceeds %D Criteria

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

CCV - Modified 03/05/2008
PDF File ID: 4822560
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574241-02
Instrument ID: HPMS15
Run Time: 12:37
Method: 8270D
File ID: 15M18593
Analyst: MES
QC Key: DOD4
Workgroup (AAB#): WG573853
Cal ID: HPMS15 - 21-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
1,4-Dioxane		5000	4360	ug/L	0.267	12.8	20	

* Exceeds %D Criteria

CCC Calibration Check Compounds
SPCC System Performance Check Compounds

CCV - Modified 03/05/2008
PDF File ID: 4822560
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16061026
Instrument ID: HPMS15
Workgroup (AAB#): WG573853

ICAL CCV Number: WG573501-02
CAL ID: HPMS15-21-JUN-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1
WG573501-02	NA	NA	39637
Upper Limit	NA	NA	79274
Lower Limit	NA	NA	19819
L16061026-04	1.00	01	55282
L16061026-05	40.0	DL01	46495
L16061026-07	1.00	01	44830
WG573299-01	1.00	01	57608
WG573299-02	1.00	01	51597
WG573299-03	1.00	01	54701

IS-1 - 1,4-Dichlorobenzene-d4

Underline = Response outside limits

INTERNAL_STD_ICAL - Modified 03/06/2008
PDF File ID: 4822561
Report generated 06/29/2016 08:21

Microbac Laboratories Inc.

INTERNAL STANDARD RETENTION TIME SUMMARY
(COMPARED TO MIDPOINT OF ICAL)

Login Number: L16061026
Instrument ID: HPMS15
Workgroup (AAB#): WG573853

ICAL CCV Number: WG573501-02
CAL ID: HPMS15-21-JUN-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1
WG573501-02	NA	NA	6.93
Upper Limit	NA	NA	7.43
Lower Limit	NA	NA	6.43
L16061026-04	1.00	01	6.929
L16061026-05	40.0	DL01	6.896
L16061026-07	1.00	01	6.929
WG573299-01	1.00	01	6.929
WG573299-02	1.00	01	6.929
WG573299-03	1.00	01	6.929

IS-1 - 1,4-Dichlorobenzene-d4

Underline = Response outside limits

INTERNAL_STD_RT_ICAL - Modified 03/06/2008
PDF File ID: 4828846
Report generated: 06/29/2016 08:21

2.2.1.3 Sample Data

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18521.D
Acq On : 23 Jun 2016 18:08
Operator : MES
Sample : L16061026-04 827-DIOX
Misc : 1,1
ALS Vial : 13 Sample Multiplier: 1

Quant Time: Jun 24 08:01:29 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.929	152	55282	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.771	96	130983	5.7581	ug/L	-0.06
Spiked Amount 5.000			Recovery	= 115.20%		
4) Nitrobenzene-d5	7.431	82	7452	0.1276	ug/L	0.00
Spiked Amount 5.000			Recovery	= 2.60%		
5) 2-Fluorobiphenyl	8.919	172	1842	0.0125	ug/L	0.00
Spiked Amount 5.000			Recovery	= 0.20%		
6) p-Terphenyl-d14	12.184	244	137	0.0010	ug/L	0.00
Spiked Amount 5.000			Recovery	= 0.00%		
Target Compounds						Qvalue
2) 1,4-Dioxane	3.897	58	110340	6.5234	ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18521.D
Acq On : 23 Jun 2016 18:08
Operator : MES
Sample : L16061026-04 827-DIOX
Misc : 1,1
ALS Vial : 13 Sample Multiplier: 1

Quant Time: Jun 24 08:01:29 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 6.52 ug/L
RT: 3.897 min Scan# 116
Delta R.T. 0.035 min
Lab File: 15M18521.D
Acq: 23 Jun 2016 18:08

Tgt Ion: 58 Resp: 110340

Ion	Ratio	Lower	Upper
58	100		
88	151.6	103.1	240.7



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062716\
Data File : 15M18600.D
Acq On : 27 Jun 2016 15:30
Operator : MES
Sample : L16061026-05 40X
Misc : 1,40
ALS Vial : 9 Sample Multiplier: 40

Quant Time: Jun 27 17:44:13 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.896	152	46495	1.0000	ug/mL	#-0.04
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.665	96	84849	177.3975	ug/L	-0.16
Spiked Amount 5.000			Recovery	= 3548.00%		
4) Nitrobenzene-d5	7.404	82	94	0.0766	ug/L	-0.03
Spiked Amount 5.000			Recovery	= 1.60%		
5) 2-Fluorobiphenyl	8.889	172	85	0.0274	ug/L	-0.03
Spiked Amount 5.000			Recovery	= 0.60%		
6) p-Terphenyl-d14	12.103	244	3	0.0010	ug/L	-0.08
Spiked Amount 5.000			Recovery	= 0.00%		
Target Compounds						Qvalue
2) 1,4-Dioxane	3.799	58	78397	5.5109	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062716\
Data File : 15M18600.D
Acq On : 27 Jun 2016 15:30
Operator : MES
Sample : L16061026-05 40X
Misc : 1,40
ALS Vial : 9 Sample Multiplier: 40

Quant Time: Jun 27 17:44:13 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 5.51 ug/L
RT: 3.799 min Scan# 102
Delta R.T. -0.063 min
Lab File: 15M18600.D
Acq: 27 Jun 2016 15:30

Tgt Ion: 58 Resp: 78397

Ion	Ratio	Lower	Upper
58	100		
88	154.3	103.1	240.7



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18523.D
Acq On : 23 Jun 2016 18:54
Operator : MES
Sample : L16061026-07 827-DIOX
Misc : 1,1
ALS Vial : 15 Sample Multiplier: 1

Quant Time: Jun 24 08:01:35 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.929	152	44830	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.848	96	46839	2.5391	ug/L	0.02
Spiked Amount 5.000			Recovery	=	50.80%	
4) Nitrobenzene-d5	7.431	82	6116	0.1292	ug/L	0.00
Spiked Amount 5.000			Recovery	=	2.60%	
5) 2-Fluorobiphenyl	8.919	172	2445	0.0205	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.40%	
6) p-Terphenyl-d14	12.151	244	474	0.0041	ug/L	-0.03
Spiked Amount 5.000			Recovery	=	0.00%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.883	58	8049	0.5868	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18523.D
Acq On : 23 Jun 2016 18:54
Operator : MES
Sample : L16061026-07 827-DIOX
Misc : 1,1
ALS Vial : 15 Sample Multiplier: 1

Quant Time: Jun 24 08:01:35 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 0.59 ug/L
RT: 3.883 min Scan# 114
Delta R.T. 0.021 min
Lab File: 15M18523.D
Acq: 23 Jun 2016 18:54

Tgt Ion: 58 Resp: 8049

Ion	Ratio	Lower	Upper
58	100		
88	156.7	103.1	240.7



2.2.1.4 Standards Data

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18458.D
Acq On : 21 Jun 2016 9:52
Operator : MES
Sample : WG573501-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 22 08:04:39 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	6.933	152	39637	1.0000 ug/mL	# 0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.827	96	82573	5.0627 ug/L	0.00
Spiked Amount 5.000			Recovery	= 101.20%	
4) Nitrobenzene-d5	7.431	82	223381	5.3353 ug/L	0.00
Spiked Amount 5.000			Recovery	= 106.80%	
5) 2-Fluorobiphenyl	8.922	172	533101	5.0444 ug/L	0.00
Spiked Amount 5.000			Recovery	= 100.80%	
6) p-Terphenyl-d14	12.184	244	519544	5.1003 ug/L	0.00
Spiked Amount 5.000			Recovery	= 102.00%	
Target Compounds					Qvalue
2) 1,4-Dioxane	3.862	58	60883	5.0202 ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18458.D
Acq On : 21 Jun 2016 9:52
Operator : MES
Sample : WG573501-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 22 08:04:39 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18459.D
Acq On : 21 Jun 2016 10:27
Operator : MES
Sample : WG573501-03 10 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 22 08:04:42 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	43874	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.834	96	186767	10.3452	ug/L	0.00
Spiked Amount 5.000			Recovery = 207.00%			
4) Nitrobenzene-d5	7.431	82	527549	11.3834	ug/L	0.00
Spiked Amount 5.000			Recovery = 227.60%			
5) 2-Fluorobiphenyl	8.926	172	1199799	10.2565	ug/L	0.00
Spiked Amount 5.000			Recovery = 205.20%			
6) p-Terphenyl-d14	12.240	244	1225328	10.8672	ug/L	0.06
Spiked Amount 5.000			Recovery = 217.40%			
Target Compounds						Qvalue
2) 1,4-Dioxane	3.869	58	138820	10.3412	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18459.D
Acq On : 21 Jun 2016 10:27
Operator : MES
Sample : WG573501-03 10 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 22 08:04:42 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18460.D
Acq On : 21 Jun 2016 10:49
Operator : MES
Sample : WG573501-04 7.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 22 08:04:45 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	6.933	152	44299	1.0000 ug/mL	# 0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.834	96	138833	7.6163 ug/L	0.00
Spiked Amount 5.000			Recovery	= 152.40%	
4) Nitrobenzene-d5	7.431	82	389810	8.3306 ug/L	0.00
Spiked Amount 5.000			Recovery	= 166.60%	
5) 2-Fluorobiphenyl	8.923	172	909050	7.6965 ug/L	0.00
Spiked Amount 5.000			Recovery	= 154.00%	
6) p-Terphenyl-d14	12.203	244	921380	8.0931 ug/L	0.02
Spiked Amount 5.000			Recovery	= 161.80%	
Target Compounds					Qvalue
2) 1,4-Dioxane	3.876	58	103182	7.6127 ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18460.D
Acq On : 21 Jun 2016 10:49
Operator : MES
Sample : WG573501-04 7.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 22 08:04:45 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18461.D
Acq On : 21 Jun 2016 11:12
Operator : MES
Sample : WG573501-05 2.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jun 22 08:04:48 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	6.933	152	35478	1.0000 ug/mL	# 0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.827	96	35630	2.4406 ug/L	0.00
Spiked Amount 5.000			Recovery	= 48.80%	
4) Nitrobenzene-d5	7.431	82	88869	2.3714 ug/L	0.00
Spiked Amount 5.000			Recovery	= 47.40%	
5) 2-Fluorobiphenyl	8.923	172	232500	2.4579 ug/L	0.00
Spiked Amount 5.000			Recovery	= 49.20%	
6) p-Terphenyl-d14	12.199	244	218191	2.3930 ug/L	0.02
Spiked Amount 5.000			Recovery	= 47.80%	
Target Compounds					Qvalue
2) 1,4-Dioxane	3.862	58	26333	2.4259 ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18461.D
Acq On : 21 Jun 2016 11:12
Operator : MES
Sample : WG573501-05 2.5 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jun 22 08:04:48 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18462.D
Acq On : 21 Jun 2016 11:34
Operator : MES
Sample : WG573501-06 1 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jun 22 08:04:51 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	44672	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.841	96	17660	0.9607	ug/L	0.01
Spiked Amount 5.000			Recovery	=	19.20%	
4) Nitrobenzene-d5	7.431	82	41409	0.8776	ug/L	0.00
Spiked Amount 5.000			Recovery	=	17.60%	
5) 2-Fluorobiphenyl	8.922	172	114491	0.9612	ug/L	0.00
Spiked Amount 5.000			Recovery	=	19.20%	
6) p-Terphenyl-d14	12.203	244	105713	0.9208	ug/L	0.02
Spiked Amount 5.000			Recovery	=	18.40%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.876	58	13250	0.9694	ug/L	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18462.D
Acq On : 21 Jun 2016 11:34
Operator : MES
Sample : WG573501-06 1 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jun 22 08:04:51 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18463.D
Acq On : 21 Jun 2016 11:57
Operator : MES
Sample : WG573501-07 0.4 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jun 22 08:04:54 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	45840	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.841	96	7629	0.4045	ug/L	0.01
Spiked Amount 5.000			Recovery	=	8.00%	
4) Nitrobenzene-d5	7.431	82	16604	0.3429	ug/L	0.00
Spiked Amount 5.000			Recovery	=	6.80%	
5) 2-Fluorobiphenyl	8.923	172	48617	0.3978	ug/L	0.00
Spiked Amount 5.000			Recovery	=	8.00%	
6) p-Terphenyl-d14	12.210	244	44097	0.3743	ug/L	0.03
Spiked Amount 5.000			Recovery	=	7.40%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.876	58	5650	0.4028	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18463.D
Acq On : 21 Jun 2016 11:57
Operator : MES
Sample : WG573501-07 0.4 PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jun 22 08:04:54 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18464.D
Acq On : 21 Jun 2016 12:20
Operator : MES
Sample : WG573501-08 5PPM 1,4-DIOXANE ALT SRC
Misc : 1,1 STD74214
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 22 08:04:57 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.933	152	41944	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.792	96	22	0.0013	ug/L	-0.04
Spiked Amount 5.000			Recovery	=	0.00%	
4) Nitrobenzene-d5	7.436	82	56	0.0013	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.00%	
5) 2-Fluorobiphenyl	8.926	172	23	0.0002	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.00%	
6) p-Terphenyl-d14	12.214	244	128	0.0012	ug/L	0.03
Spiked Amount 5.000			Recovery	=	0.00%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.869	58	64750	5.0454	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18464.D
Acq On : 21 Jun 2016 12:20
Operator : MES
Sample : WG573501-08 5PPM 1,4-DIOXANE ALT SRC
Misc : 1,1 STD74214
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 22 08:04:57 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18464.D
Acq On : 21 Jun 2016 12:20
Operator : MES
Sample : WG573501-08 5PPM 1,4-DIOXANE ALT SRC
Misc : 1,1 STD74214
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jun 22 08:04:57 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2		1,4-Dioxane	5.000	5.045	-0.9	106	0.00
3	S	1,4-Dioxane-d8	5.000	0.001	100.0#	0	-0.04
4	S	Nitrobenzene-d5	5.000	0.001	100.0#	0	0.00
5	S	2-Fluorobiphenyl	5.000	0.000	100.0#	0	0.00
6	S	p-Terphenyl-d14	5.000	0.001	100.0#	0	0.03

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18510.D
Acq On : 23 Jun 2016 13:59
Operator : MES
Sample : WG573879-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 23 14:15:55 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	6.929	152	50565	1.0000 ug/mL	# 0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.848	96	96203	4.6237 ug/L	0.02
Spiked Amount 5.000			Recovery	= 92.40%	
4) Nitrobenzene-d5	7.426	82	280398	5.2498 ug/L	0.00
Spiked Amount 5.000			Recovery	= 105.00%	
5) 2-Fluorobiphenyl	8.919	172	664952	4.9322 ug/L	0.00
Spiked Amount 5.000			Recovery	= 98.60%	
6) p-Terphenyl-d14	12.155	244	665166	5.1186 ug/L	-0.03
Spiked Amount 5.000			Recovery	= 102.40%	
Target Compounds					Qvalue
2) 1,4-Dioxane	3.883	58	70119	4.5322 ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18510.D
Acq On : 23 Jun 2016 13:59
Operator : MES
Sample : WG573879-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 23 14:15:55 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18510.D
Acq On : 23 Jun 2016 13:59
Operator : MES
Sample : WG573879-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 23 14:15:55 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.306	0.277	9.5	115	0.02
3 S	1,4-Dioxane-d8	0.411	0.381	7.3	117	0.02
4 S	Nitrobenzene-d5	1.056	1.109	-5.0	126	0.00
5 S	2-Fluorobiphenyl	2.666	2.630	1.4	125	0.00
6 S	p-Terphenyl-d14	2.570	2.631	-2.4	128	-0.03

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18510.D
Acq On : 23 Jun 2016 13:59
Operator : MES
Sample : WG573879-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 23 14:15:55 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	5.000	4.532	9.4	115	0.02
3 S	1,4-Dioxane-d8	5.000	4.624	7.5	117	0.02
4 S	Nitrobenzene-d5	5.000	5.250	-5.0	126	0.00
5 S	2-Fluorobiphenyl	5.000	4.932	1.4	125	0.00
6 S	p-Terphenyl-d14	5.000	5.119	-2.4	128	-0.03

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062716\
Data File : 15M18593.D
Acq On : 27 Jun 2016 12:37
Operator : MES
Sample : WG574241-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 27 13:07:36 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	42974	1.0000	ug/mL	#-0.04
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.749	96	78711	4.4512	ug/L	-0.08
Spiked Amount 5.000			Recovery	=	89.00%	
4) Nitrobenzene-d5	7.393	82	235288	5.1833	ug/L	-0.04
Spiked Amount 5.000			Recovery	=	103.60%	
5) 2-Fluorobiphenyl	8.886	172	551093	4.8097	ug/L	-0.04
Spiked Amount 5.000			Recovery	=	96.20%	
6) p-Terphenyl-d14	12.170	244	561144	5.0809	ug/L	-0.01
Spiked Amount 5.000			Recovery	=	101.60%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.785	58	57358	4.3623	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062716\
Data File : 15M18593.D
Acq On : 27 Jun 2016 12:37
Operator : MES
Sample : WG574241-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 27 13:07:36 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062716\
Data File : 15M18593.D
Acq On : 27 Jun 2016 12:37
Operator : MES
Sample : WG574241-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 27 13:07:36 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	-0.04
2	1,4-Dioxane	0.306	0.267	12.7	94	-0.08
3 S	1,4-Dioxane-d8	0.411	0.366	10.9	95	-0.08
4 S	Nitrobenzene-d5	1.056	1.095	-3.7	105	-0.04
5 S	2-Fluorobiphenyl	2.666	2.565	3.8	103	-0.04
6 S	p-Terphenyl-d14	2.570	2.612	-1.6	108	-0.01

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Evaluate Continuing Calibration Report

Data Path : C:\msdchem\1\data\062716\
Data File : 15M18593.D
Acq On : 27 Jun 2016 12:37
Operator : MES
Sample : WG574241-02 5PPM 1,4-DIOXANE STD
Misc : 1,1 STD74784
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 27 13:07:36 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	-0.04
2	1,4-Dioxane	5.000	4.362	12.8	94	-0.08
3 S	1,4-Dioxane-d8	5.000	4.451	11.0	95	-0.08
4 S	Nitrobenzene-d5	5.000	5.183	-3.7	105	-0.04
5 S	2-Fluorobiphenyl	5.000	4.810	3.8	103	-0.04
6 S	p-Terphenyl-d14	5.000	5.081	-1.6	108	-0.01

(#) = Out of Range SPCC's out = 0 CCC's out = 0

2.2.1.5 Raw QC Data

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\msdchem\1\methods\DFTPPQ.M
Title : DFTPP
Last Update : Mon Apr 27 13:16:06 2015



AutoFind: Scans 234, 235, 236; Background Corrected with Scan 228

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	38.8	34488	PASS
68	69	0.00	2	1.4	500	PASS
69	198	0.00	100	41.4	36833	PASS
70	69	0.00	2	0.5	172	PASS
127	198	40	60	51.9	46131	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	88880	PASS
199	198	5	9	6.7	5918	PASS
275	198	10	30	24.9	22090	PASS
365	198	1	100	2.7	2358	PASS
441	443	0.01	100	75.7	9586	PASS
442	198	40	100	73.4	65261	PASS
443	442	17	23	19.4	12659	PASS

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 21 09:47:12 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18457.D\data.ms

(1) Pentachlorophenol

7.054min (+ 0.105) 0.00 ug/ml

response 30738

Ion	Exp%	Act%
265.80	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 21 09:47:12 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





(2) Benzidine

8.502min (+ 0.008) 0.00 ug/ml

response 532770

Ion	Exp%	Act%
184.10	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062116\
Data File : 15M18457.D
Acq On : 21 Jun 2016 9:34
Operator : MES
Sample : WG573501-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 21 09:47:12 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18457.D\data.ms

(3) DDT

9.486min (-0.108) 0.00 ug/ml

response 180187

Ion	Exp%	Act%
235.00	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18509.D
Acq On : 23 Jun 2016 13:42
Operator : MES
Sample : WG573879-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\msdchem\1\methods\DFTPPQ.M
Title : DFTPP
Last Update : Mon Apr 27 13:16:06 2015





AutoFind: Scans 233, 234, 235; Background Corrected with Scan 227

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	37.0	43111	PASS
68	69	0.00	2	1.1	506	PASS
69	198	0.00	100	39.4	45795	PASS
70	69	0.00	2	0.2	72	PASS
127	198	40	60	51.3	59752	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	116364	PASS
199	198	5	9	7.1	8249	PASS
275	198	10	30	24.9	28968	PASS
365	198	1	100	2.8	3278	PASS
441	443	0.01	100	79.2	13093	PASS
442	198	40	100	74.3	86464	PASS
443	442	17	23	19.1	16541	PASS

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18509.D
Acq On : 23 Jun 2016 13:42
Operator : MES
Sample : WG573879-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 23 13:53:21 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18509.D\data.ms

(1) Pentachlorophenol

7.048min (+ 0.099) 0.00 ug/ml

response 32081

Ion	Exp%	Act%
265.80	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18509.D
Acq On : 23 Jun 2016 13:42
Operator : MES
Sample : WG573879-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 23 13:53:21 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18509.D\data.ms

(2) Benzidine

8.496min (+ 0.002) 0.00 ug/ml

response 452358

Ion	Exp%	Act%
184.10	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18509.D
Acq On : 23 Jun 2016 13:42
Operator : MES
Sample : WG573879-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 23 13:53:21 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





(3) DDT

9.594min (-9.594) 0.00 ug/ml

response 0

Ion	Exp%	Act%
235.00	100.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062716\
Data File : 15M18592.D
Acq On : 27 Jun 2016 12:20
Operator : MES
Sample : WG574241-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\msdchem\1\methods\DFTPPQ.M
Title : DFTPP
Last Update : Mon Apr 27 13:16:06 2015





AutoFind: Scans 227, 228, 229; Background Corrected with Scan 221

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	33.2	31989	PASS
68	69	0.00	2	1.6	580	PASS
69	198	0.00	100	37.4	35954	PASS
70	69	0.00	2	0.6	222	PASS
127	198	40	60	50.4	48459	PASS
197	198	0.00	1	0.4	409	PASS
198	198	100	100	100.0	96232	PASS
199	198	5	9	6.6	6399	PASS
275	198	10	30	25.2	24236	PASS
365	198	1	100	2.5	2410	PASS
441	443	0.01	100	75.7	10371	PASS
442	198	40	100	73.9	71133	PASS
443	442	17	23	19.3	13708	PASS

Data Path : C:\msdchem\1\data\062716\
Data File : 15M18592.D
Acq On : 27 Jun 2016 12:20
Operator : MES
Sample : WG574241-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 27 12:31:57 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





(1) Pentachlorophenol

7.014min (+ 0.065) 0.00 ug/ml

response 37276

Ion	Exp%	Act%
265.80	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062716\
Data File : 15M18592.D
Acq On : 27 Jun 2016 12:20
Operator : MES
Sample : WG574241-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 27 12:31:57 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18592.D\data.ms

(2) Benzidine

8.461min (-0.033) 0.00 ug/ml

response 225697

Ion	Exp%	Act%
184.10	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : C:\msdchem\1\data\062716\
Data File : 15M18592.D
Acq On : 27 Jun 2016 12:20
Operator : MES
Sample : WG574241-01 5PPM DFTPP STD
Misc : 1,1 STD75016
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 27 12:31:57 2016
Quant Method : C:\msdchem\1\methods\DFTPPQ.M
Quant Title : DFTPP
QLast Update : Mon Apr 27 13:16:06 2015
Response via : Initial Calibration





TIC: 15M18592.D\data.ms

(3) DDT

9.428min (-0.166) 0.00 ug/ml mint

response 273417

Ion	Exp%	Act%
235.00	100.00	100.00
0.00	0.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18511.D
Acq On : 23 Jun 2016 14:22
Operator : MES
Sample : WG573299-01 BLANK 6/21
Misc : 1,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 23 15:52:24 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.929	152	57608	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.862	96	33985	1.4337	ug/L	0.04
Spiked Amount 5.000			Recovery = 28.60%			
4) Nitrobenzene-d5	7.426	82	1290209	21.2028	ug/L	0.00
Spiked Amount 5.000			Recovery = 424.00%			
5) 2-Fluorobiphenyl	8.919	172	2634137	17.1496	ug/L	0.00
Spiked Amount 5.000			Recovery = 343.00%			
6) p-Terphenyl-d14	12.162	244	4420616	29.8588	ug/L	-0.02
Spiked Amount 5.000			Recovery = 597.20%			

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18511.D
Acq On : 23 Jun 2016 14:22
Operator : MES
Sample : WG573299-01 BLANK 6/21
Misc : 1,1
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 23 15:52:24 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18512.D
Acq On : 23 Jun 2016 14:45
Operator : MES
Sample : WG573299-02 LCS 6/21
Misc : 1,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 23 15:52:38 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.929	152	51597	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.862	96	40084	1.8880	ug/L	0.04
Spiked Amount 5.000			Recovery	=	37.80%	
4) Nitrobenzene-d5	7.431	82	2991	0.0549	ug/L	0.00
Spiked Amount 5.000			Recovery	=	1.00%	
5) 2-Fluorobiphenyl	8.919	172	1228	0.0089	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.20%	
6) p-Terphenyl-d14	12.159	244	1342	0.0101	ug/L	-0.03
Spiked Amount 5.000			Recovery	=	0.20%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.897	58	29708	1.8818	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18512.D
Acq On : 23 Jun 2016 14:45
Operator : MES
Sample : WG573299-02 LCS 6/21
Misc : 1,1
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jun 23 15:52:38 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18513.D
Acq On : 23 Jun 2016 15:07
Operator : MES
Sample : WG573299-03 LCSDUP 6/21
Misc : 1,1
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jun 23 15:52:42 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.929	152	54701	1.0000	ug/mL	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.862	96	52635	2.3384	ug/L	0.04
Spiked Amount 5.000			Recovery	=	46.80%	
4) Nitrobenzene-d5	7.431	82	5866	0.1015	ug/L	0.00
Spiked Amount 5.000			Recovery	=	2.00%	
5) 2-Fluorobiphenyl	8.919	172	4684	0.0321	ug/L	0.00
Spiked Amount 5.000			Recovery	=	0.60%	
6) p-Terphenyl-d14	12.151	244	1266	0.0090	ug/L	-0.03
Spiked Amount 5.000			Recovery	=	0.20%	
Target Compounds						Qvalue
2) 1,4-Dioxane	3.897	58	38804	2.3185	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : C:\msdchem\1\data\062316\
Data File : 15M18513.D
Acq On : 23 Jun 2016 15:07
Operator : MES
Sample : WG573299-03 LCSDUP 6/21
Misc : 1,1
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jun 23 15:52:42 2016
Quant Method : C:\msdchem\1\methods\DIOXANE_D8.M
Quant Title : OVD MSS01 SIM 1,4-dioxane ICAL 062116
QLast Update : Tue Jun 21 12:51:17 2016
Response via : Initial Calibration



2.2 General Chromatography Data

2.2.2 6850 LC/MS Data

2.2.2.1 Summary Data

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-01	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	18WW16-061616	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 19:40
Collect Date:	06/16/2016 07:55	**Dilution:**	1	**File ID:**	1LM.LM35503
Sample Tag:	01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	5.76		0.400	0.200	0.100

Certificate of Analysis

Sample #:	L16061026-03	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	102-061616	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 01:12
Collect Date:	06/16/2016 08:40	**Dilution:**	100	**File ID:**	1LM.LM35479
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	107		40.0	20.0	10.0

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	129-061616	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 01:31
Collect Date:	06/16/2016 09:05	**Dilution:**	10000	**File ID:**	1LM.LM35480
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	8950		4000	2000	1000

Certificate of Analysis

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Sample #:	L16061026-05	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	120-061616	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 19:59
Collect Date:	06/16/2016 09:25	**Dilution:**	10000	**File ID:**	1LM.LM35504
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	32700		4000	2000	1000

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	LCMS1
Client ID:	123-061616	**Prep Method:**	6850	**Prep Date:**	06/24/2016 19:00
Matrix:	Water	**Analytical Method:**	6850	**Cal Date:**	05/03/2016 17:18
Workgroup #:	WG574101	**Analyst:**	JWR	**Run Date:**	06/25/2016 20:18
Collect Date:	06/16/2016 09:55	**Dilution:**	50	**File ID:**	1LM.LM35505
Sample Tag:	DL01	**Units:**	ug/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Perchlorate	14797-73-0	105		20.0	10.0	5.00

2.2.2.2 QC Summary Data

Example Calculation 6850 - Perchlorate

Concentration from Linear Regression

Step 1: Retrieve Curve Data From Plot, y = mx + b

y = response ratio = response of analyte / response of internal standard (IS) = Rx/Ristd

x = amount ratio = concentration analyte/concentration internal standard (IS) = Cx / Cistd

m = slope from curve (1.45)

b = intercept from curve (-0.00242)

y = 1.45x + -0.00242

Step 2: Substitute the value for y

where y = 12600/226000 = 0.055752

Step 3: Solve for x

x = (y - b)/m = 0.0040119

Step 4: Solve for analyte concentration Cx

Cx = (Cis)(x) = (5 ug/L)(0.040119) = 0.200594 ug/L

Example Calculation - Water:

Slope from curve, m:	1.45
Intercept from curve, b:	-0.00242
Response of analyte, Rx:	12600
Response of Internal Standard , Ristd:	226000
Concentration of IS, Cistd (ug/L):	5.00
Response Ratio:	0.05575
Amount Ratio:	0.04012
Analyte Concentration, Cx (ug/L) :	0.200594

Example Calculation - Soil:

Analyte Concentration, Cx (ug/L):	0.20059
Amount of soil extracted (g):	5.00
Final volume of extract (mL):	50.00
Percent solids (Pct wt.)	100
Concentration in soil (ug/kg):	2.005938

Perchlorate Conductivity Check
(perchlorate1)

Conductivity Probe Calibration Check: 1380 /1410 µs/cm

Working MCT Level: 10,000 µs/cm

Sample	Conductivity (µs/cm)	Pretreatment or Dilution Needed
WG574101-01 MCT	9,990.	
-02 Blank	0.33	
-03 LCS	0.48	
-04 LCS2	0.51	
L16060823-01	9,120.	
-03	7,440.	
-04	1,943.	
L16060927-01	750.	
-02	652.	
-03	446.	
-04	162.7	
L16061026-01	6,910.	
-03	238.	
-04	965.	
-05	1,526.	
-07	220.	
L16061142-01	165.3	
-02	81.1	
-04	83.1	
-06	728.	
-07	74.2	
L16061144-01	118.6	
-02	118.7	
-03	110.1	

Analyst: John Richards

Date/Time: 06/27/16 / 20:00

DCN#119317

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 050316_JWR.TXT
Analyst1: JWR Analyst2: NA
Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254
Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG567013 (soils)

Internal STD: COA18071 Surrogate STD: NA Calibration STD STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 MS/MSD STD: STD75512

Comments: ICAL WG567320 : Alternate Source STD75512
Analytical Column : RPPX 5um (250x4.6mm)
K'Prime S/N RPPX250-02115

Samples L16041363(-05 and -10) were analyzed at dilutions based on their pre-run screen results.

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	1LM.LM34686	WG567320-01 CCB	1	1		05/03/16 15:06
2	1LM.LM34687	WG567320-02 STD (0.1 ug/L)	1	1	STD75510	05/03/16 15:25
3	1LM.LM34688	WG567320-03 STD (0.2 ug/L)	1	1	STD75510	05/03/16 15:43
4	1LM.LM34689	WG567320-04 STD (0.5 ug/L)	1	1	STD75510	05/03/16 16:02
5	1LM.LM34690	WG567320-05 STD (1.0 ug/L)	1	1	STD75510	05/03/16 16:21
6	1LM.LM34691	WG567320-06 STD (2.0 ug/L)	1	1	STD75510	05/03/16 16:40
7	1LM.LM34692	WG567320-07 STD (5.0 ug/L)	1	1	STD75510	05/03/16 16:59
8	1LM.LM34693	WG567320-08 STD (10 ug/L)	1	1	STD75510	05/03/16 17:18
9	1LM.LM34694	WG567320-09 SSCV (1.0 ug/L)	1	1	STD75512	05/03/16 17:37
10	1LM.LM34695	WG567321-01 CCB	1	1		05/03/16 17:56
11	1LM.LM34696	WG567321-02 CCV (1.0ug/L)	1	1	STD75510	05/03/16 18:15
12	1LM.LM34697	WG567013-07 MRL (2.0ug/kg)	7	1	STD75510	05/03/16 18:34
13	1LM.LM34698	WG567013-01 MCT (2.0ug/kg)	7	1	STD75512	05/03/16 18:53
14	1LM.LM34699	WG567013-02 BLANK	7	1		05/03/16 19:12
15	1LM.LM34700	WG567013-03 LCS (2.0ug/kg)	7	1	STD75512	05/03/16 19:31
16	1LM.LM34701	L16041363-07 RS	7	1		05/03/16 19:50
17	1LM.LM34702	L16041363-08 MS	7	1	STD75512	05/03/16 20:09
18	1LM.LM34703	L16041363-09 MSD	7	1	STD75512	05/03/16 20:28
19	1LM.LM34704	L16041363-01	7	1		05/03/16 20:46
20	1LM.LM34705	L16041363-02	7	1		05/03/16 21:05
21	1LM.LM34706	L16041363-03	7	1		05/03/16 21:24
22	1LM.LM34707	L16041363-04	7	1		05/03/16 21:43
23	1LM.LM34708	WG567321-03 CCV (1.0ug/L)	1	1	STD75510	05/03/16 22:02
24	1LM.LM34709	WG567013-08 MRL (2.0ug/kg)	7	1	STD75510	05/03/16 22:21
25	1LM.LM34710	WG567321-04 CCB	1	1		05/03/16 22:40
26	1LM.LM34711	L16041363-05 (5x)	7	5		05/03/16 22:59
27	1LM.LM34712	L16041363-06	7	1		05/03/16 23:18
28	1LM.LM34713	L16041363-10 (5x)	7	5		05/03/16 23:37
29	1LM.LM34714	WG567321-05 CCV (1.0ug/L)	1	1	STD75510	05/03/16 23:56
30	1LM.LM34715	WG567013-09 MRL (2.0ug/kg)	7	1	STD75510	05/04/16 00:15
31	1LM.LM34716	WG567321-06 CCB	1	1		05/04/16 00:34

Comments

Page: 1 Approved: 05-MAY-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 050316_JWR.TXT

Analyst1: JWR Analyst2: NA

Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254

Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG567013 (soils)

Internal STD: COA18071 Surrogate STD: NA STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 STD75512

Comments

Seq.	Rerun	Dil.	Reason	Analytes
17				
	L16041363-08 MS : The MS %Rec is 129%. This is above the advisory limit of 120%. The parent sample to this MS had responses that passed the ion-ratio criteria, but had a quantified value below the method's detection limit, resulting in an assigned value of zero. If the quantified value for the parent sample were used in the %Rec calculation, the MS %Rec would be 93.9%.			
18				
	L16041363-09 MSD : The MSD %Rec is 131%. This is above the advisory limit of 120%. The parent sample to this MSD had responses that passed the ion-ratio criteria, but had a quantified value below the method's detection limit, resulting in an assigned value of zero. If the quantified value for the parent sample were used in the %Rec calculation, the MSD %Rec would be 95.4%.			

Page: 2 Approved: 05-MAY-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 062416_JWR.TXT
Analyst1: JWR Analyst2: NA
Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254
Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG574101 (waters)

Internal STD: STD18071 Surrogate STD: NA Calibration STD STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 MS/MSD STD: NA

Comments: Samples initially analyzed at dilutions were done so based on their historical results. Samples tagged with (NR) did not conform to their historical results. These samples were reanalyzed on the end of the sequence as neat or at dilutions based on the results of their initial analysis.

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
1	1LM.LM35462	WG574102-01 CCB	1	1		06/24/16 19:50
2	1LM.LM35463	WG574102-02 CCV (1.0ug/L)	1	1	STD75510	06/24/16 20:09
3	1LM.LM35464	WG574101-05 MRL (0.2ug/L)	1	1	STD75510	06/24/16 20:28
4	1LM.LM35465	WG574101-01 MCT (0.2ug/L)	1	1	STD75512	06/24/16 20:47
5	1LM.LM35466	WG574101-02 BLANK	1	1		06/24/16 21:06
6	1LM.LM35467	WG574101-03 LCS (0.2ug/L)	1	1	STD75512	06/24/16 21:25
7	1LM.LM35468	L16060823-01	1	1		06/24/16 21:44
8	1LM.LM35469	L16060823-03	1	1		06/24/16 22:03
9	1LM.LM35470	L16060823-04 (10x)	1	10		06/24/16 22:22
10	1LM.LM35471	L16060927-01 (100x) (NR)	1	100		06/24/16 22:41
11	1LM.LM35472	L16060927-02 (100x) (NR)	1	100		06/24/16 23:00
12	1LM.LM35473	L16060927-03	1	1		06/24/16 23:18
13	1LM.LM35474	L16060927-04	1	1		06/24/16 23:37
14	1LM.LM35475	WG574102-03 CCV (1.0ug/L)	1	1	STD75510	06/24/16 23:56
15	1LM.LM35476	WG574101-06 MRL (0.2ug/L)	1	1	STD75510	06/25/16 00:15
16	1LM.LM35477	WG574102-04 CCB	1	1		06/25/16 00:34
17	1LM.LM35478	L16061026-01 (100x) (NR)	1	100		06/25/16 00:53
18	1LM.LM35479	L16061026-03 (100x)	1	100		06/25/16 01:12
19	1LM.LM35480	L16061026-04 (10,000x)	1	10000		06/25/16 01:31
20	1LM.LM35481	L16061026-05 (100,000x) (NR)	1	100000		06/25/16 01:50
21	1LM.LM35482	L16061026-07 (10x) (NR)	1	10		06/25/16 02:09
22	1LM.LM35483	L16061142-01	1	1		06/25/16 02:28
23	1LM.LM35484	L16061142-02	1	1		06/25/16 02:47
24	1LM.LM35485	L16061142-04	1	1		06/25/16 03:06
25	1LM.LM35486	L16061142-06 (10x) (NR)	1	10		06/25/16 03:25
26	1LM.LM35487	L16061142-07	1	1		06/25/16 03:44
27	1LM.LM35488	WG574102-05 CCV (1.0ug/L)	1	1	STD75510	06/25/16 04:03
28	1LM.LM35489	WG574101-07 MRL (0.2ug/L)	1	1	STD75510	06/25/16 04:22
29	1LM.LM35490	WG574102-06 CCB	1	1		06/25/16 04:40
30	1LM.LM35491	WG574101-04 LCS2 (0.2ug/L)	1	1	STD75512	06/25/16 04:59
31	1LM.LM35492	L16061144-01	1	1		06/25/16 05:18
32	1LM.LM35493	L16061144-02	1	1		06/25/16 05:37
33	1LM.LM35494	L16061144-03 (10x) (NR)	1	10		06/25/16 05:56

Page: 1 Approved: 28-JUN-16

Microbac Laboratories Inc.

Instrument Run Log

Instrument: LCMS1 Dataset: 062416_JWR.TXT
Analyst1: JWR Analyst2: NA
Method: 6850 SOP: HPLC06 Rev: 8

Maintenance Log ID: Syringe Filter Lot#: 151125254
Eluent ID#:

Column 1 ID: KP-RPPX250 Column 2 ID: NA

Workgroups: Analytical WG574101 (waters)

Internal STD: STD18071 Surrogate STD: NA STD75510 (05/03/2016)

CCV STD: STD75510 LCS STD: STD75512 NA

Seq.	File ID	Sample Information	Mat	Dil	Reference	Date/Time
34	1LM.LM35495	WG574102-07 CCV (1.0ug/L)	1	1	STD75510	06/25/16 06:15
35	1LM.LM35496	WG574101-08 MRL (0.2ug/L)	1	1	STD75510	06/25/16 06:34
36	1LM.LM35497	WG574102-08 CCB	1	1		06/25/16 06:53
37	1LM.LM35498	WG574102-09 CCV (1.0ug/L)	1	1	STD75510	06/25/16 18:06
38	1LM.LM35499	WG574101-09 MRL (0.2ug/L)	1	1	STD75510	06/25/16 18:25
39	1LM.LM35500	WG574102-10 CCB	1	1		06/25/16 18:44
40	1LM.LM35501	L16060927-01 (RR Neat)	1	1		06/25/16 19:02
41	1LM.LM35502	L16060927-02 (RR 1000x)	1	1000		06/25/16 19:21
42	1LM.LM35503	L16061026-01 (RR Neat)	1	1		06/25/16 19:40
43	1LM.LM35504	L16061026-05 (RR 10,000x)	1	10000		06/25/16 19:59
44	1LM.LM35505	L16061026-07 (RR 50x)	1	50		06/25/16 20:18
45	1LM.LM35506	L16061142-06 (RR Neat)	1	1		06/25/16 20:37
46	1LM.LM35507	L16061144-03 (RR Neat)	1	1		06/25/16 20:56
47	1LM.LM35508	WG574102-11 CCV (1.0ug/L)	1	1	STD75510	06/25/16 21:15
48	1LM.LM35509	WG574101-10 MRL (0.2ug/L)	1	1	STD75510	06/25/16 21:34
49	1LM.LM35510	WG574102-12 CCB	1	1		06/25/16 21:53

Comments

Seq.	Rerun	Dil.	Reason	Analytes

Page: 2 Approved: 28-JUN-16

Microbac Laboratories Inc.

Data Checklist

Date: 03-MAY-2016
Analyst: JWR
Analyst: NA
Method: 6850
Instrument: LCMS1
Curve Workgroup: WG567320
Runlog ID: 74891
Analytical Workgroups: L16041363 (SOILS)

ANALYTICAL	
System Performance Check	NA
DFTPP (GCMS)	NA
Endrin/DDT breakdown (8081/GCMS)	NA
Pentachlorophenol/benzidine tailing (GCMS)	NA
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	X
Average RF	NA
Linear regression or higher order curve	X
Alternate source standard (ICV) % Difference	X
Continuing Calibration (CCV)	X
% D/% Drift	X
Minimum response factors (GCMS)	X
Continuing calibration blank (CCB) (IC/LCMS)	X
Limit of quantitation verification (LOQV) (LCMS)	X
Special standards	NA
Blanks	X
TCL hits	ND
Surrogate recoveries	NA
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	NA
MS/MSD/Sample duplicates	X
Recoveries	X
%RPD	X
Interference check sample (ICS) (LCMS)	MCT
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	NA
Surrogate recoveries	NA
Internal standard areas (MS)	X
Library searches (GCMS)	NA
Calculations & correct factors	X
Compounds above calibration range	NA
Reruns	NA
Manual integrations	NA
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	
Check for completeness	X
Primary Reviewer	JWR
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	WTD

Primary Reviewer:
04-MAY-2016
John Richards

Secondary Reviewer:
05-MAY-2016
Wadei D

CHECKLIST1 - Modified 03/05/2008
Generated: MAY-05-2016 16:23:46

Microbac Laboratories Inc.

Data Checklist

Date: 24-JUN-2016
Analyst: JWR
Analyst: NA
Method: 6850
Instrument: LCMS1
Curve Workgroup: NA
Runlog ID: 75960
Analytical Workgroups: L16060823, 0927, 1026, 1142, 1144

ANALYTICAL	
System Performance Check	NA
DFTPP (GCMS)	NA
Endrin/DDT breakdown (8081/GCMS)	NA
Pentachlorophenol/benzidine tailing (GCMS)	NA
Eluent check (IC)/system pressure (HPLC)	NA
Window standard (FID)	NA
Initial Calibration	NA
Average RF	NA
Linear regression or higher order curve	NA
Alternate source standard (ICV) % Difference	NA
Continuing Calibration (CCV)	X
% D/% Drift	X
Minimum response factors (GCMS)	X
Continuing calibration blank (CCB) (IC/LCMS)	X
Limit of quantitation verification (LOQV) (LCMS)	X
Special standards	NA
Blanks	X
TCL hits	ND
Surrogate recoveries	NA
LCS/LCSD (Laboratory Control Sample)	X
Recoveries	X
Surrogate recoveries	NA
MS/MSD/Sample duplicates	NA
Recoveries	NA
%RPD	NA
Interference check sample (ICS) (LCMS)	MCT
Samples	X
TCL hits	X
Mass spectra (MS/HPLC)/2nd column confirmations (ECD/FID/HPLC)	NA
Surrogate recoveries	NA
Internal standard areas (MS)	X
Library searches (GCMS)	NA
Calculations & correct factors	X
Compounds above calibration range	X
Reruns	X
Manual integrations	MCT-ONLY
Project/client specific requirements	X
REPORTING	
Upload batch form	X
KOBRA workgroup data/forms/bench sheets	X
Case narratives	
Check for completeness	X
Primary Reviewer	JWR
SUPERVISORY/SECONDARY REVIEW	
Check for compliance with method and project specific requirements	X
Check the completeness/accuracy of reported information	X
Data qualifiers	X
Secondary Reviewer	WTD

Primary Reviewer:
27-JUN-2016

Secondary Reviewer:
28-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-28-2016 14:00:55

Microbac Laboratories Inc.

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 6850 AAB#: WG574101

Login Number: L16061026

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18WW16-061616	01	06/16/16					06/24/2016	8.5	28		06/25/16	1	28	
102-061616	03	06/16/16					06/24/2016	8.4	28		06/25/16	.3	28	
129-061616	04	06/16/16					06/24/2016	8.4	28		06/25/16	.3	28	
120-061616	05	06/16/16					06/24/2016	8.4	28		06/25/16	1	28	
123-061616	07	06/16/16					06/24/2016	8.4	28		06/25/16	1.1	28	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4827778
Report generated 06/28/2016 14:17

METHOD BLANK SUMMARY

Login Number: L16061026
Blank File ID: 1LM.LM35466
Prep Date: 06/24/16 19:00
Analyzed Date: 06/24/16 21:06
Analyst: JWR

Work Group: WG574101
Blank Sample ID: WG574101-02
Instrument ID: LCMS1
Method: 6850

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
QCMRL	WG574101-05	1LM.LM35464	06/24/16 20:28	01
MCT	WG574101-01	1LM.LM35465	06/24/16 20:47	01
LCS	WG574101-03	1LM.LM35467	06/24/16 21:25	01
QCMRL	WG574101-06	1LM.LM35476	06/25/16 00:15	01
102-061616	L16061026-03	1LM.LM35479	06/25/16 01:12	DL01
129-061616	L16061026-04	1LM.LM35480	06/25/16 01:31	DL01
QCMRL	WG574101-07	1LM.LM35489	06/25/16 04:22	01
LCS2	WG574101-04	1LM.LM35491	06/25/16 04:59	01
QCMRL	WG574101-08	1LM.LM35496	06/25/16 06:34	01
QCMRL	WG574101-09	1LM.LM35499	06/25/16 18:25	01
18WW16-061616	L16061026-01	1LM.LM35503	06/25/16 19:40	01
120-061616	L16061026-05	1LM.LM35504	06/25/16 19:59	DL01
123-061616	L16061026-07	1LM.LM35505	06/25/16 20:18	DL01
QCMRL	WG574101-10	1LM.LM35509	06/25/16 21:34	01

Report Name: BLANK_SUMMARY
PDF File ID: 4827779
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16061026 Prep Date: 06/24/16 19:00 Sample ID: WG574101-02
Instrument ID: LCMS1 Run Date: 06/24/16 21:06 Prep Method: 6850
File ID: 1LM.LM35466 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Perchlorate	0.100	0.400	0.100	1	U

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name:BLANK
PDF ID: 4827780
28-JUN-2016 14:17

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16061026 Analyst: JWR Prep Method: 6850
Instrument ID: LCMS1 Matrix: Water Method: 6850
Workgroup (AAB#): WG574101 Units: ug/L
QC Key: DOD4 Lot #: STD75512

Sample ID: WG574101-03 LCS File ID: 1LM.LM35467 Run Date: 06/24/2016 21:25
Sample ID: WG574101-04 LCS2 File ID: 1LM.LM35491 Run Date: 06/25/2016 04:59

Analytes	LCS			LCS2			%RPD	%Rec Limits	RPD Lmt	Q
	Known	Found	% REC	Known	Found	% REC				
Perchlorate	0.200	0.208	104	0.200	0.213	107	2.38	80 - 120	15	

LCS_LCS2 - Modified 03/06/2008
PDF File ID: 4827781
Report generated: 06/28/2016 14:17

Microbac Laboratories Inc.

INITIAL CALIBRATION SUMMARY

Login Number: L16061026
Analytical Method: 6850
ICAL Workgroup: WG567320

Instrument ID: LCMS1
Initial Calibration Date: 03-MAY-16 17:18
Column ID: F

Analyte		AVG RF	% RSD	LINEAR (R)	QUAD (R^2)
Perchlorate		1.699	4.81	1.00000	

R = Correlation coefficient; 0.995 minimum
R^2 = Coefficient of determination; 0.99 minimum

INT_CAL - Modified 03/06/2008
PDF File ID: 4828717
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16061026

Analytical Method: 6850

Instrument ID: LCMS1

Initial Calibration Date: 03-MAY-16 17:18

Column ID: F

Analyte	WG567320-02			WG567320-03			WG567320-04		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
Perchlorate	0.100	17900.0000	1.792	0.200	34100.0000	1.718	0.500	82200.0000	1.637

INT_CAL - Modified 03/06/2008
PDF File ID: 4828717
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16061026

Analytical Method: 6850

Instrument ID: LCMS1

Initial Calibration Date: 03-MAY-16 17:18

Column ID: F

Analyte	WG567320-05			WG567320-06			WG567320-07		
	CONC	RESP	RF	CONC	RESP	RF	CONC	RESP	RF
Perchlorate	1.00	168000.000	1.697	2.00	330000.000	1.672	5.00	810000.000	1.695

INT_CAL - Modified 03/06/2008
PDF File ID: 4828717
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

INITIAL CALIBRATION DATA

Login Number: L16061026

Analytical Method: 6850

Instrument ID: LCMS1

Initial Calibration Date: 03-MAY-16 17:18

Column ID: F

Analyte	WG567320-08		
	CONC	RESP	RF
Perchlorate	10.0	1530000.00	1.680

INT_CAL - Modified 03/06/2008
PDF File ID: 4828717
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.
ALTERNATE SOURCE CALIBRATION REPORT

Login Number: L16061026 Run Date: 05/03/2016 Sample ID: WG567320-09
Instrument ID: LCMS1 Run Time: 17:37 Method: 6850
File ID: 1LM.LM34694 Analyst: JWR QC Key: DOD4
ICal Workgroup: WG567320 Cal ID: LCMS1 - 03-MAY-16

Analyte		Expected	Found	Units	RF	%D	UCL	Q
Perchlorate		1.00	0.985	ug/L	1.66	1.50	15	

* Exceeds %D Limit

ALT - Modified 09/06/2007
Version 1.5 PDF File ID: 4828718
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/24/2016
Sample ID: WG574102-01
Instrument ID: LCMS1
Run Time: 19:50
Method: 6850
File ID: 1LM.LM35462
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827784
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/25/2016
Sample ID: WG574102-04
Instrument ID: LCMS1
Run Time: 00:34
Method: 6850
File ID: 1LM.LM35477
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827784
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/25/2016
Sample ID: WG574102-06
Instrument ID: LCMS1
Run Time: 04:40
Method: 6850
File ID: 1LM.LM35490
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827784
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/25/2016
Sample ID: WG574102-08
Instrument ID: LCMS1
Run Time: 06:53
Method: 6850
File ID: 1LM.LM35497
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827784
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/25/2016
Sample ID: WG574102-10
Instrument ID: LCMS1
Run Time: 18:44
Method: 6850
File ID: 1LM.LM35500
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827784
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/25/2016
Sample ID: WG574102-12
Instrument ID: LCMS1
Run Time: 21:53
Method: 6850
File ID: 1LM.LM35510
Analyst: JWR
Units: ug/L
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Perchlorate	0.100	0.400	0.100	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4827784
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026 Run Date: 06/24/2016 Sample ID: WG574102-02
Instrument ID: LCMS1 Run Time: 20:09 Method: 6850
File ID: 1LM.LM35463 Analyst: JWR QC Key: DOD4
Workgroup (AAB#): WG574101 Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.01	ug/L	1.70	1.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827783
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/24/2016
Sample ID: WG574102-03
Instrument ID: LCMS1
Run Time: 23:56
Method: 6850
File ID: 1LM.LM35475
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.01	ug/L	1.70	1.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827783
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026 Run Date: 06/25/2016 Sample ID: WG574102-05
Instrument ID: LCMS1 Run Time: 04:03 Method: 6850
File ID: 1LM.LM35488 Analyst: JWR QC Key: DOD4
Workgroup (AAB#): WG574101 Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.04	ug/L	1.75	4.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827783
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/25/2016
Sample ID: WG574102-07
Instrument ID: LCMS1
Run Time: 06:15
Method: 6850
File ID: 1LM.LM35495
Analyst: JWR
QC Key: DOD4
Workgroup (AAB#): WG574101
Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.02	ug/L	1.72	2.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827783
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026 Run Date: 06/25/2016 Sample ID: WG574102-09
Instrument ID: LCMS1 Run Time: 18:06 Method: 6850
File ID: 1LM.LM35498 Analyst: JWR QC Key: DOD4
Workgroup (AAB#): WG574101 Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.04	ug/L	1.76	4.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827783
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026 Run Date: 06/25/2016 Sample ID: WG574102-11
Instrument ID: LCMS1 Run Time: 21:15 Method: 6850
File ID: 1LM.LM35508 Analyst: JWR QC Key: DOD4
Workgroup (AAB#): WG574101 Cal ID: LCMS1 - 03-MAY-16
Matrix: WATER

Analyte		Expected	Found	UNITS	RF	%D	UCL	Q
Perchlorate		1.00	1.02	ug/L	1.73	2.00	15	

* Exceeds %D Criteria

CCV - Modified 03/05/2008
PDF File ID: 4827783
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16061026 Run Date: 06/24/2016 Sample ID: WG574101-05
Instrument ID: LCMS1 Run Time: 20:28 Prep Method: 6850
File ID: 1LM.LM35464 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.206	103	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827782
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16061026 Run Date: 06/25/2016 Sample ID: WG574101-06
Instrument ID: LCMS1 Run Time: 00:15 Prep Method: 6850
File ID: 1LM.LM35476 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.198	99.0	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827782
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16061026 Run Date: 06/25/2016 Sample ID: WG574101-07
Instrument ID: LCMS1 Run Time: 04:22 Prep Method: 6850
File ID: 1LM.LM35489 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.211	106	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827782
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16061026 Run Date: 06/25/2016 Sample ID: WG574101-08
Instrument ID: LCMS1 Run Time: 06:34 Prep Method: 6850
File ID: 1LM.LM35496 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.201	101	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827782
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16061026 Run Date: 06/25/2016 Sample ID: WG574101-09
Instrument ID: LCMS1 Run Time: 18:25 Prep Method: 6850
File ID: 1LM.LM35499 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.207	104	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827782
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

QCMRL SAMPLE

Login Number: L16061026 Run Date: 06/25/2016 Sample ID: WG574101-10
Instrument ID: LCMS1 Run Time: 21:34 Prep Method: 6850
File ID: 1LM.LM35509 Analyst: JWR Method: 6850
Workgroup (AAB#): WG574101 Matrix: Water Units: ug/L
Contract #: Cal ID: LCMS1-03-MAY-16

Analytes	Expected	Found	% Rec	Limits	Q
Perchlorate	0.200	0.206	103	70 - 130	

QCMRL - Modified 03/06/2007
PDF File ID: 4827782
Report generated 06/28/2016 14:17

Microbac Laboratories Inc.

INTERNAL STANDARD AREA SUMMARY
(COMPARED TO AVERAGE OF ICAL)

Login Number: L16061026
Instrument ID: LCMS1
Workgroup (AAB#): WG574101

ICAL CCV Number: WG567320-05
CAL ID: LCMS1-03-MAY-16
Matrix: WATER

Sample Number	Dilution	Tag	IS-1
WG567320	NA	NA	489000
Upper Limit	NA	NA	733500
Lower Limit	NA	NA	244500
L16061026-01	1.00	01	359000
L16061026-03	100	DL01	498000
L16061026-04	10000	DL01	491000
L16061026-05	10000	DL01	457000
L16061026-07	50.0	DL01	490000
WG574101-02	1.00	01	443000
WG574101-03	1.00	01	444000
WG574101-04	1.00	01	526000

IS-1 - O18LP

Underline = Response outside limits

INTERNAL_STD_AVG_ICAL - Modified 03/10/2010
PDF File ID: 4827785
Report generated 06/28/2016 14:17

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	L16061026-01
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35503
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 19:40	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	697000	245000	2.84	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	L16061026-03
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35479
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 01:12	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	180000	63800	2.82	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	L16061026-04
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35480
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 01:31	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	148000	50500	2.93	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	L16061026-05
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35504
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 19:59	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	504000	174000	2.90	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	L16061026-07
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35505
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 20:18	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	348000	120000	2.90	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG567320-02
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34687
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 15:25	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	17900	6950	2.58	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG567320-03
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34688
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 15:43	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	34100	11900	2.87	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG567320-04
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34689
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:02	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	82200	29400	2.80	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG567320-05
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34690
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:21	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	168000	56600	2.97	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG567320-06
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34691
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:40	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	330000	108000	3.06	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG567320-07
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34692
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 16:59	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	810000	269000	3.01	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG567320-08
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34693
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 17:18	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	1530000	512000	2.99	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG567320-09
Instrument:	LCMS1	**Prep Date:**		**File ID:**	1LM.LM34694
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	05/03/2016 17:37	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	169000	56300	3.00	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	WG574101-01
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35465
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 20:47	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	28100	10000	2.81	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	WG574101-02
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35466
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 21:06	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	WG574101-03
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35467
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 21:25	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	31600	11800	2.68	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	WG574101-04
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35491
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:59	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	38300	14000	2.74	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	WG574101-05
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35464
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 20:28	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	31600	12100	2.61	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	WG574101-06
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35476
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 00:15	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	35000	13100	2.67	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	WG574101-07
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35489
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:22	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	38900	13400	2.90	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	WG574101-08
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35496
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 06:34	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	38000	14100	2.70	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	WG574101-09
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35499
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 18:25	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	32900	11500	2.86	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**	6850	**Samplenum:**	WG574101-10
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:00	**File ID:**	1LM.LM35509
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 21:34	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	37900	14000	2.71	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-01
Instrument:	LCMS1	**Prep Date:**	06/24/2016 19:50	**File ID:**	1LM.LM35462
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 19:50	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-02
Instrument:	LCMS1	**Prep Date:**	06/24/2016 20:09	**File ID:**	1LM.LM35463
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 20:09	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	152000	55500	2.74	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-03
Instrument:	LCMS1	**Prep Date:**	06/24/2016 23:56	**File ID:**	1LM.LM35475
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/24/2016 23:56	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	170000	57500	2.96	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-04
Instrument:	LCMS1	**Prep Date:**	06/25/2016 00:34	**File ID:**	1LM.LM35477
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 00:34	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	824	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-05
Instrument:	LCMS1	**Prep Date:**	06/25/2016 04:03	**File ID:**	1LM.LM35488
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:03	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	179000	60200	2.97	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-06
Instrument:	LCMS1	**Prep Date:**	06/25/2016 04:40	**File ID:**	1LM.LM35490
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 04:40	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-07
Instrument:	LCMS1	**Prep Date:**	06/25/2016 06:15	**File ID:**	1LM.LM35495
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 06:15	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	189000	63400	2.98	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-08
Instrument:	LCMS1	**Prep Date:**	06/25/2016 06:53	**File ID:**	1LM.LM35497
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 06:53	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	518	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-09
Instrument:	LCMS1	**Prep Date:**	06/25/2016 18:06	**File ID:**	1LM.LM35498
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 18:06	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	161000	56700	2.84	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-10
Instrument:	LCMS1	**Prep Date:**	06/25/2016 18:44	**File ID:**	1LM.LM35500
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 18:44	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-11
Instrument:	LCMS1	**Prep Date:**	06/25/2016 21:15	**File ID:**	1LM.LM35508
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 21:15	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	176000	59300	2.97	2.3	3.8	

Perchlorate Ion Ratios

Microbac Laboratories Inc.

Login #:	L16061026	**Prep Method:**		**Samplenum:**	WG574102-12
Instrument:	LCMS1	**Prep Date:**	06/25/2016 21:53	**File ID:**	1LM.LM35510
Analyst:	JWR	**Anal Method:**	6850	**Matrix:**	Water
Worknum:	WG574101	**Analysis Date:**	06/25/2016 21:53	**Units:**	ug/L

Analyte	Res #1	Res #2	Ratio	Lower	Upper	Q
PERCHLORATE	0.000	0.000	0.000	2.3	3.8	*

2.2.2.3 Sample Data

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:53 PM

Data File	LM35503.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 7:40:50 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16061026-01 (RR Neat)	**Injection Vial**	30.00
Data File	LM35503.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 7:40:50 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16061026-01	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	3.590e+05	9.15	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	6.970e+05	9.16	N/A	5.76
Perchlorate conf	2.450e+05	9.14	N/A	6.10

O18LP (Internal Standard)

RT (Exp. RT): 9.15(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:53 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.16 (9.84) min Calculated conc: 5.76 ng/ml Area Ratio: 1.94 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.14 (9.84) min Calculated conc: 6.10 ng/ml Area Ratio: 0.683 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:44 PM

Data File	LM35479.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 1:12:33 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16061026-03 (100x)	**Injection Vial**	16.00
Data File	LM35479.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 1:12:33 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16061026-03	**Dilution Factor**	1.00
Sample Comment	1,100 (Hist)	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.980e+05	9.60	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.800e+05	9.62	N/A	1.07
Perchlorate conf	6.380e+04	9.60	N/A	1.13

O18LP (Internal Standard)

RT (Exp. RT): 9.60(9.84) min
Concentration: 5.00 ug/L
Sample Type: (Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:44 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.62 (9.84) min
Calculated conc:	1.07 ng/ml
Area Ratio:	0.362
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.60 (9.84) min
Calculated conc:	1.13 ng/ml
Area Ratio:	0.128
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:45 PM

Data File	LM35480.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 1:31:29 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16061026-04 (10,000x)	**Injection Vial**	17.00
Data File	LM35480.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 1:31:29 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16061026-04	**Dilution Factor**	1.00
Sample Comment	1,10000 (Hist)	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.910e+05	9.61	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.480e+05	9.62	N/A	0.895
Perchlorate conf	5.050e+04	9.60	N/A	0.90

O18LP (Internal Standard)

RT (Exp. RT): 9.61(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:45 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.62 (9.84) min
Calculated conc:	0.895 ng/ml
Area Ratio:	0.302
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.60 (9.84) min
Calculated conc:	0.90 ng/ml
Area Ratio:	0.103
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:54 PM

Data File	LM35504.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 7:59:46 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16061026-05 (RR 10,000x)	**Injection Vial**	31.00
Data File	LM35504.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 7:59:46 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16061026-05	**Dilution Factor**	1.00
Sample Comment	110,000.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.570e+05	9.62	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	5.040e+05	9.64	N/A	3.27
Perchlorate conf	1.740e+05	9.62	N/A	3.38

O18LP (Internal Standard)

RT (Exp. RT): 9.62(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.64 (9.84) min Calculated conc: 3.27 ng/ml Area Ratio: 1.101 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.62 (9.84) min Calculated conc: 3.38 ng/ml Area Ratio: 0.38 Sample Type: (Unknown)	

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:54 PM

Data File	LM35505.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 8:18:43 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	L16061026-07 (RR 50x)	**Injection Vial**	32.00
Data File	LM35505.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 8:18:43 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	L16061026-07	**Dilution Factor**	1.00
Sample Comment	150.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.900e+05	9.62	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.480e+05	9.64	N/A	2.10
Perchlorate conf	1.200e+05	9.62	N/A	2.18

O18LP (Internal Standard)

RT (Exp. RT): 9.62(9.84) min
Concentration: 5.00 ug/L
Sample Type: (Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:54 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.64 (9.84) min Calculated conc: 2.10 ng/ml Area Ratio: 0.709 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.62 (9.84) min Calculated conc: 2.18 ng/ml Area Ratio: 0.245 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

2.2.2.4 Standards Data

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Data File	LM34686.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-01 CCB	**Injection Vial**	1.00
Data File	LM34686.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-01	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.020e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min
Concentration: 5.00 ug/L
Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:20 PM

Analyte Name: **Perchlorate**
Internal Standard: O18LP

Data File	LM34686.wiff	**Result Table**	050316_JWR.rdb
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Regression Equation: y = 1.68 x + 0.00128 (r = 1.0000)

Expected Concentration	Number of Values	Mean Calculated Concentration	% Accuracy	Std. Deviation	%CV
0.10	1	0.10	102.8	N/A	N/A
0.20	1	0.20	100.3	N/A	N/A
0.50	1	0.48	96.6	N/A	N/A
1.00	1	1.01	100.5	N/A	N/A
2.00	1	1.99	99.3	N/A	N/A
5.00	1	5.04	100.7	N/A	N/A
10.00	1	9.99	99.9	N/A	N/A



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:20 PM

Analyte Name: **Perchlorate conf**
Internal Standard: O18LP

Data File	LM34686.wiff	**Result Table**	050316_JWR.rdb
Acquisition Date	5/3/2016 3:06:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Regression Equation: y = 0.558 x + 0.00228 (r = 0.9999)

Expected Concentration	Number of Values	Mean Calculated Concentration	% Accuracy	Std. Deviation	%CV
0.10	1	0.10	104.3	N/A	N/A
0.20	1	0.19	96.8	N/A	N/A
0.50	1	0.50	100.6	N/A	N/A
1.00	1	1.00	100.5	N/A	N/A
2.00	1	1.94	97.2	N/A	N/A
5.00	1	5.02	100.4	N/A	N/A
10.00	1	10.03	100.3	N/A	N/A



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Data File	LM34687.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 3:25:04 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-02 STD (0.1 ug/L)	**Injection Vial**	2.00
Data File	LM34687.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 3:25:04 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-02	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.000e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.790e+04	9.85	0.10	0.103
Perchlorate conf	6.950e+03	9.83	0.10	0.104

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:36 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.103 ng/ml
Area Ratio:	0.036
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	0.104 ng/ml
Area Ratio:	0.014
Sample Type:	(Standard)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34688.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 3:43:59 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-03 STD (0.2 ug/L)	**Injection Vial**	3.00
Data File	LM34688.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 3:43:59 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-03	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.970e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.410e+04	9.85	0.20	0.201
Perchlorate conf	1.190e+04	9.84	0.20	0.194

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.201 ng/ml
Area Ratio:	0.069
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	0.194 ng/ml
Area Ratio:	0.024
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34689.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:02:52 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-04 STD (0.5 ug/L)	**Injection Vial**	4.00
Data File	LM34689.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:02:52 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-04	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.020e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	8.220e+04	9.85	0.50	0.483
Perchlorate conf	2.940e+04	9.84	0.50	0.503

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.483 ng/ml
Area Ratio:	0.164
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	0.503 ng/ml
Area Ratio:	0.058
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34690.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:21:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-05 STD (1.0 ug/L)	**Injection Vial**	5.00
Data File	LM34690.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:21:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-05	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.950e+05	9.83	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.680e+05	9.84	1.00	1.01
Perchlorate conf	5.660e+04	9.83	1.00	1.00

O18LP (Internal Standard)

RT (Exp. RT): 9.83(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.339
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	1.00 ng/ml
Area Ratio:	0.114
Sample Type:	(Standard)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Data File	LM34691.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:40:45 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-06 STD (2.0 ug/L)	**Injection Vial**	6.00
Data File	LM34691.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:40:45 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-06	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.930e+05	9.83	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.300e+05	9.84	2.00	1.99
Perchlorate conf	1.080e+05	9.83	2.00	1.94

O18LP (Internal Standard)

RT (Exp. RT): 9.83(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:37 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	1.99 ng/ml
Area Ratio:	0.669
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	1.94 ng/ml
Area Ratio:	0.219
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Data File	LM34692.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 4:59:42 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-07 STD (5.0 ug/L)	**Injection Vial**	7.00
Data File	LM34692.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 4:59:42 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-07	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.780e+05	9.83	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	8.100e+05	9.84	5.00	5.04
Perchlorate conf	2.690e+05	9.83	5.00	5.02

O18LP (Internal Standard)

RT (Exp. RT): 9.83(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	5.04 ng/ml
Area Ratio:	1.695
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	5.02 ng/ml
Area Ratio:	0.562
Sample Type:	(Standard)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Data File	LM34693.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	5/3/2016 5:18:37 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-08 STD (10 ug/L)	**Injection Vial**	8.00
Data File	LM34693.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 5:18:37 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Standard
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG567320-08	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.560e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.530e+06	9.85	10.00	9.99
Perchlorate conf	5.120e+05	9.83	10.00	10.00

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Standard)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	9.99 ng/ml
Area Ratio:	3.36
Sample Type:	(Standard)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.83 (9.84) min
Calculated conc:	10.00 ng/ml
Area Ratio:	1.122
Sample Type:	(Standard)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:23 PM

Data File	LM34694.wiff	**Result Table**	050316_JWR.rdb
Acquisition Date	5/3/2016 5:37:34 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG567320-09 SSCV (1.0 ug/L)	**Injection Vial**	9.00
Data File	LM34694.wiff	**Injection Volume**	10.00
Acquisition Date	5/3/2016 5:37:34 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	050316_JWR.rdb
Sample ID	WG567320-09	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.080e+05	9.84	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.690e+05	9.85	1.00	0.985
Perchlorate conf	5.630e+04	9.84	1.00	0.973

O18LP (Internal Standard)

RT (Exp. RT): 9.84(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 5/4/2016 1:23 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.85 (9.84) min
Calculated conc:	0.985 ng/ml
Area Ratio:	0.332
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.84 (9.84) min
Calculated conc:	0.973 ng/ml
Area Ratio:	0.111
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Data File	LM35463.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 8:09:38 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-02 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35463.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 8:09:38 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-02	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.450e+05	9.34	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.520e+05	9.35	1.00	1.01
Perchlorate conf	5.550e+04	9.33	1.00	1.10

O18LP (Internal Standard)

RT (Exp. RT): 9.34(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.341
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.33 (9.84) min
Calculated conc:	1.10 ng/ml
Area Ratio:	0.125
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Data File	LM35475.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 11:56:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-03 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35475.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 11:56:49 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-03	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.980e+05	9.32	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.700e+05	9.33	1.00	1.01
Perchlorate conf	5.750e+04	9.32	1.00	1.01

O18LP (Internal Standard)

RT (Exp. RT): 9.32(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.33 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.34
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.32 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.115
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:47 PM

Data File	LM35488.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:03:05 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-05 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35488.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:03:05 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-05	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.110e+05	9.33	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.790e+05	9.34	1.00	1.04
Perchlorate conf	6.020e+04	9.32	1.00	1.03

O18LP (Internal Standard)

RT (Exp. RT): 9.33(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:47 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.34 (9.84) min
Calculated conc:	1.04 ng/ml
Area Ratio:	0.35
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.32 (9.84) min
Calculated conc:	1.03 ng/ml
Area Ratio:	0.118
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:50 PM

Data File	LM35495.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:15:37 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-07 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35495.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:15:37 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-07	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.500e+05	9.33	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.890e+05	9.35	1.00	1.02
Perchlorate conf	6.340e+04	9.33	1.00	1.01

O18LP (Internal Standard)

RT (Exp. RT): 9.33(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.02 ng/ml
Area Ratio:	0.345
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.33 (9.84) min
Calculated conc:	1.01 ng/ml
Area Ratio:	0.115
Sample Type:	(Quality Control)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Data File	LM35498.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:06:10 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-09 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35498.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:06:10 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-09	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.580e+05	9.34	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.610e+05	9.35	1.00	1.04
Perchlorate conf	5.670e+04	9.34	1.00	1.09

O18LP (Internal Standard)

RT (Exp. RT): 9.34(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.04 ng/ml
Area Ratio:	0.353
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.34 (9.84) min
Calculated conc:	1.09 ng/ml
Area Ratio:	0.124
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:55 PM

Data File	LM35508.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 9:15:29 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-11 CCV (1.0ug/L)	**Injection Vial**	3.00
Data File	LM35508.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 9:15:29 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Quality Control
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-11	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.090e+05	9.35	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	1.760e+05	9.36	1.00	1.02
Perchlorate conf	5.930e+04	9.35	1.00	1.02

O18LP (Internal Standard)

RT (Exp. RT): 9.35(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Quality Control)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:55 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.36 (9.84) min
Calculated conc:	1.02 ng/ml
Area Ratio:	0.345
Sample Type:	(Quality Control)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	1.02 ng/ml
Area Ratio:	0.117
Sample Type:	(Quality Control)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Data File	LM35464.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 8:28:33 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-05 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35464.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 8:28:33 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-05	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.490e+05	9.39	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.160e+04	9.40	N/A	0.206
Perchlorate conf	1.210e+04	9.37	N/A	0.222

O18LP (Internal Standard)

RT (Exp. RT): 9.39(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.40 (9.84) min
Calculated conc:	0.206 ng/ml
Area Ratio:	0.07
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.37 (9.84) min
Calculated conc:	0.222 ng/ml
Area Ratio:	0.027
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Data File	LM35476.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 12:15:45 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-06 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35476.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 12:15:45 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-06	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.160e+05	9.36	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.500e+04	9.37	N/A	0.198
Perchlorate conf	1.310e+04	9.35	N/A	0.207

O18LP (Internal Standard)

RT (Exp. RT): 9.36(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.37 (9.84) min
Calculated conc:	0.198 ng/ml
Area Ratio:	0.068
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.35 (9.84) min
Calculated conc:	0.207 ng/ml
Area Ratio:	0.025
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Data File	LM35489.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:22:00 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-07 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35489.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:22:00 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-07	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.380e+05	9.37	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.890e+04	9.39	N/A	0.211
Perchlorate conf	1.340e+04	9.38	N/A	0.202

O18LP (Internal Standard)

RT (Exp. RT): 9.37(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.39 (9.84) min Calculated conc: 0.211 ng/ml Area Ratio: 0.072 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.38 (9.84) min Calculated conc: 0.202 ng/ml Area Ratio: 0.025 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:50 PM

Data File	LM35496.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:34:33 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-08 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35496.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:34:33 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-08	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.520e+05	9.38	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.800e+04	9.39	N/A	0.201
Perchlorate conf	1.410e+04	9.38	N/A	0.208

O18LP (Internal Standard)

RT (Exp. RT): 9.38(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.39 (9.84) min
Calculated conc:	0.201 ng/ml
Area Ratio:	0.069
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.38 (9.84) min
Calculated conc:	0.208 ng/ml
Area Ratio:	0.026
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Data File	LM35499.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:25:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-09 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35499.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:25:05 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-09	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.640e+05	9.39	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.290e+04	9.40	N/A	0.207
Perchlorate conf	1.150e+04	9.38	N/A	0.201

O18LP (Internal Standard)

RT (Exp. RT): 9.39(9.84) min
Concentration: 5.00 ug/L
Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.40 (9.84) min
Calculated conc:	0.207 ng/ml
Area Ratio:	0.071
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.38 (9.84) min
Calculated conc:	0.201 ng/ml
Area Ratio:	0.025
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:55 PM

Data File	LM35509.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 9:34:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-10 MRL (0.2ug/L)	**Injection Vial**	2.00
Data File	LM35509.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 9:34:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-10	**Dilution Factor**	1.00
Sample Comment	1,1 STD75510	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.370e+05	9.39	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.790e+04	9.40	N/A	0.206
Perchlorate conf	1.400e+04	9.39	N/A	0.214

O18LP (Internal Standard)

RT (Exp. RT): 9.39(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:55 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 9.40 (9.84) min Calculated conc: 0.206 ng/ml Area Ratio: 0.071 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.39 (9.84) min Calculated conc: 0.214 ng/ml Area Ratio: 0.026 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Data File	LM35462.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 7:50:39 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-01 CCB	**Injection Vial**	1.00
Data File	LM35462.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 7:50:39 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-01	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.420e+05	9.58	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.58(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:38 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:43 PM

Data File	LM35477.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 12:34:42 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-04 CCB	**Injection Vial**	1.00
Data File	LM35477.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 12:34:42 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-04	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.240e+05	9.57	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	8.240e+02	9.53	N/A	< 0

O18LP (Internal Standard)

RT (Exp. RT): 9.57(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.53 (9.84) min
Calculated conc:	< 0 ng/ml
Area Ratio:	0.002
Sample Type:	(Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Data File	LM35490.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:40:56 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-06 CCB	**Injection Vial**	1.00
Data File	LM35490.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:40:56 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-06	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.420e+05	9.59	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.59(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Data File	LM35497.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:53:28 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-08 CCB	**Injection Vial**	1.00
Data File	LM35497.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:53:28 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-08	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.550e+05	9.59	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	5.180e+02	9.63	N/A	< 0

O18LP (Internal Standard)

RT (Exp. RT): 9.59(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:51 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 9.63 (9.84) min Calculated conc: < 0 ng/ml Area Ratio: 0.001 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Data File	LM35500.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 6:44:00 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-10 CCB	**Injection Vial**	1.00
Data File	LM35500.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 6:44:00 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-10	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.840e+05	9.61	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.61(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:52 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:56 PM

Data File	LM35510.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 9:53:20 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574102-12 CCB	**Injection Vial**	1.00
Data File	LM35510.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 9:53:20 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574102-12	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.380e+05	9.61	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.61(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:56 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	0.00 (9.84) min
Calculated conc:	No Peak ng/ml
Area Ratio:	0.00
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Data File	LM35465.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 8:47:28 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-01 MCT (0.2ug/L)	**Injection Vial**	4.00
Data File	LM35465.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 8:47:28 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-01	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.320e+05	8.95	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	2.810e+04	8.97	N/A	0.19
Perchlorate conf	1.000e+04	8.95	N/A	0.188

O18LP (Internal Standard)

RT (Exp. RT): 8.95(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:39 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	8.97 (9.84) min
Calculated conc:	0.19 ng/ml
Area Ratio:	0.065
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	8.95 (9.84) min
Calculated conc:	0.188 ng/ml
Area Ratio:	0.023
Sample Type:	(Unknown)



Printing Time: 4:34:28 PM
Printing Date: Saturday, June 25, 2016

Sample Name: "WG574101-01 MCT (0.2ug/L)" Sample ID: "WG574101-01" File: "LM35465.wiff"
Peak Name: "Perchlorate" Mass(es): "98.8/83.3 amu"
Comment: "1,1 STD75512" Annotation: ""

Sample Index:	1
Sample Type:	Unknown
Concentration:	N/A
Calculated Conc:	0.219 ug/L
Acq. Date:	6/24/2016
Acq. Time:	8:47:28 PM
Modified:	No
Proc. Algorithm:	Analyst Classic
Bunching Factor:	2
Noise Threshold:	18.48 cps
Area Threshold:	92.40 cps
Num. Smooths:	5
Exp. Width:	0.20
Sep. Height:	0.01
Exp. Peak Ratio:	5.00
Exp. Adj. Ratio:	4.00
Exp. Val. Ratio:	3.00
RT Window:	30.000 sec
Expected RT:	9.84 min
Use Relative RT:	Yes
Int. Type:	Valley
Retention Time:	8.97 min
Area:	3.23e+004 counts
Height:	1951.523 cps
Start Time:	8.526 min
End Time:	9.417 min



Collected by: N/A
Electronic Signature: no
Operator: lcmsl

Printing Time: 4:36:56 PM
Printing Date: Saturday, June 25, 2016

Sample Name: "WG574101-01 MCT (0.2ug/L)" Sample ID: "WG574101-01" File: "LM35465.wiff"
Peak Name: "Perchlorate" Mass(es): "98.8/83.3 amu"
Comment: "1,1 STD75512" Annotation: ""

Sample Index: 1
Sample Type: Unknown
Concentration: N/A
Calculated Conc: 0.190 ug/L
Acq. Date: 6/24/2016
Acq. Time: 8:47:28 PM

Modified: Yes
Proc. Algorithm: Analyst Classic
Smoothing Factor: 2
Noise Threshold: 18.48 cps
Area Threshold: 92.40 cps
Num. Smooths: 5
Sep. Width: 0.20
Sep. Height: 0.01
Exp. Peak Ratio: 5.00
Exp. Adj. Ratio: 4.00
Exp. Val. Ratio: 3.00
RT Window: 30.000 sec
Expected RT: 9.84 min
Use Relative RT: Yes

Int. Type: Manual
Retention Time: 8.97 min
Area: 2.81e+004 counts
Height: 1872.249 cps
Start Time: 8.587 min
End Time: 9.417 min



#4
JWR 06/25/16

✓ WTD
06-28-16

Collected by: N/A
Electronic Signature: no
Operator: lcms1

2.2.2.5 Raw QC Data

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Data File	LM35466.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 9:06:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-02 BLANK	**Injection Vial**	5.00
Data File	LM35466.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 9:06:25 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-02	**Dilution Factor**	1.00
Sample Comment	11.00	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.430e+05	9.50	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	0.000e+00	0.00	N/A	No Peak
Perchlorate conf	0.000e+00	0.00	N/A	No Peak

O18LP (Internal Standard)

RT (Exp. RT): 9.50(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Perchlorate (98.8/83.3 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	
Perchlorate conf (100.8/85.2 amu) RT (Exp. RT): 0.00 (9.84) min Calculated conc: No Peak ng/ml Area Ratio: 0.00 Sample Type: (Unknown)	

s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Data File	LM35467.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/24/2016 9:25:22 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-03 LCS (0.2ug/L)	**Injection Vial**	6.00
Data File	LM35467.wiff	**Injection Volume**	10.00
Acquisition Date	6/24/2016 9:25:22 PM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-03	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	4.440e+05	9.53	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.160e+04	9.54	N/A	0.208
Perchlorate conf	1.180e+04	9.53	N/A	0.217

O18LP (Internal Standard)

RT (Exp. RT): 9.53(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:40 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.54 (9.84) min
Calculated conc:	0.208 ng/ml
Area Ratio:	0.071
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.53 (9.84) min
Calculated conc:	0.217 ng/ml
Area Ratio:	0.027
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Data File	LM35491.wiff	**Result Table**	062416_JWR.rdb
Acquisition Date	6/25/2016 4:59:54 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Instrument Name**	API 4000
Project	Perchlorate\2009_07_22		

Sample Name	WG574101-04 LCS2 (0.2ug/L)	**Injection Vial**	7.00
Data File	LM35491.wiff	**Injection Volume**	10.00
Acquisition Date	6/25/2016 4:59:54 AM	**Algorithm Used**	Analyst Classic
Acquisition Method	062911.dam	**Sample Type**	Unknown
Instrument Name	API 4000	**Result Table**	062416_JWR.rdb
Sample ID	WG574101-04	**Dilution Factor**	1.00
Sample Comment	1,1 STD75512	**Weight to Volume**	0.00

Internal Standard	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
O18LP	5.260e+05	9.57	5.00	-

Target Analyte	Area (cps)	RT (min)	Target conc. (ug/L)	Calc. Conc. (ug/L)
Perchlorate	3.830e+04	9.59	N/A	0.213
Perchlorate conf	1.400e+04	9.57	N/A	0.218

O18LP (Internal Standard)

RT (Exp. RT): 9.57(9.84) min

Concentration: 5.00 ug/L

Sample Type: (Unknown)



s.dataFile Page 1 of 2

Applied Biosystems | MDS SCIEX

Created with Analyst Reporter
Printed: 6/27/2016 7:48 PM

Perchlorate (98.8/83.3 amu)

RT (Exp. RT):	9.59 (9.84) min
Calculated conc:	0.213 ng/ml
Area Ratio:	0.073
Sample Type:	(Unknown)



Perchlorate conf (100.8/85.2 amu)

RT (Exp. RT):	9.57 (9.84) min
Calculated conc:	0.218 ng/ml
Area Ratio:	0.027
Sample Type:	(Unknown)



s.dataFile Page 2 of 2

2.3 Metals Data

2.3.1 Metals I C P Data

2.3.1.1 Summary Data

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 11:42
Collect Date:	06/16/2016 07:55	**Dilution:**	1	**File ID:**	T3.062716.114220
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	0.331		0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Potassium, Total	7440-09-7	13.0		2.00	1.00	0.500
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 13:00
Collect Date:	06/16/2016 07:55	**Dilution:**	20	**File ID:**	T3.062716.130008
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Calcium, Total	7440-70-2	195		10.0	5.00	2.50
Iron, Total	7439-89-6	21.5		4.00	2.00	1.00
Magnesium, Total	7439-95-4	146		20.0	10.0	5.00
Sodium, Total	7440-23-5	587		20.0	10.0	5.00

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	129-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 11:46
Collect Date:	06/16/2016 09:05	**Dilution:**	1	**File ID:**	T3.062716.114614
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	0.644		0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Calcium, Total	7440-70-2	13.2		0.500	0.250	0.125
Iron, Total	7439-89-6	1.18		0.200	0.100	0.0500
Magnesium, Total	7439-95-4	10.4		1.00	0.500	0.250
Potassium, Total	7440-09-7	0.716	J	2.00	1.00	0.500
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	129-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 13:03
Collect Date:	06/16/2016 09:05	**Dilution:**	20	**File ID:**	T3.062716.130358
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Sodium, Total	7440-23-5	146		20.0	10.0	5.00

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-06	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	120F-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 11:50
Collect Date:	06/16/2016 09:25	**Dilution:**	1	**File ID:**	T3.062716.115009
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	1.53		0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Calcium, Total	7440-70-2	6.89		0.500	0.250	0.125
Iron, Total	7439-89-6	2.04		0.200	0.100	0.0500
Magnesium, Total	7439-95-4	7.27		1.00	0.500	0.250
Potassium, Total	7440-09-7	1.55	J	2.00	1.00	0.500
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-06	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	120F-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 13:15
Collect Date:	06/16/2016 09:25	**Dilution:**	20	**File ID:**	T3.062716.131528
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Sodium, Total	7440-23-5	303		20.0	10.0	5.00

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	ICP-THERMO3
Client ID:	123-061616	**Prep Method:**	3015	**Prep Date:**	06/22/2016 07:53
Matrix:	Water	**Analytical Method:**	6010C	**Cal Date:**	06/27/2016 10:26
Workgroup #:	WG574214	**Analyst:**	JYH	**Run Date:**	06/27/2016 11:54
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	T3.062716.115400
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Aluminum, Total	7429-90-5	1.42		0.200	0.100	0.0500
Beryllium, Total	7440-41-7	0.0100	U	0.0200	0.0100	0.00500
Calcium, Total	7440-70-2	8.16		0.500	0.250	0.125
Iron, Total	7439-89-6	1.66		0.200	0.100	0.0500
Magnesium, Total	7439-95-4	4.83		1.00	0.500	0.250
Potassium, Total	7440-09-7	0.713	J	2.00	1.00	0.500
Selenium, Total	7782-49-2	0.0100	U	0.0200	0.0100	0.00500
Sodium, Total	7440-23-5	26.0		1.00	0.500	0.250

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.3.1.2 QC Summary Data

Example 6010 Calculations
Thermo Scientific iCAP

1.0 Initial Calibration (ICAL) Parameters
For a multi-point calibration, the system performs linear regression from data consisting of a blank and four standards.

2.0 Calculating the concentration (C) of an element in water using data from prep log, run log, and quantitation report (note:the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system in ug/mL (ppm)	0.1
Vf = Final volume (mL)	50
Vi = Initial volume (mL)	50
D = Dilution factor as a multiplier (10X = 10)	1
Cx = Concentration of element in ug/mL (mg/L)	0.1

3.0 Calculating the concentration (C) of an element in soil using data from prep log, run log, and quantitation report (note: the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (mg/L) (ppm)	0.1
Vf = Final volume (mL)	50
Vi = Initial weight (g)	1
D = Dilution factor as a multiplier (10X = 10)	1
Cx = Concentration of element in ug/g (mg/kg)	5

4.0 Adjusting the concentration to dry weight:

$$Cdry = \frac{Cx \times 100}{Px}$$

Where:	**Example:**
Cx = Concentration calculated as received (wet basis)	5
Px = Percent solids of sample (%wt)	80
$Cdry$ = Concentration calculated as dry weight (mg/kg)	6.25

Microbac Laboratories Inc.
Microwave Digestion Log

Workgroup: WG573656
Analyst: ERP
Spike Analyst: ERP
Run Date: 06/22/2016 07:53
Method: 3015
Balance: BAL019
Instrument: MW-4
Instrument Start: 06/22/2016 08:14

SOP: ME407 Revison 19
Spike Solution: STD76576
Spike Witness: VC
HNO3 Lot #: COA18905
HCL Lot #: COA18769
40 & 50 ML. DIGESTION TU COA18987
ICP Filters- fisher-Lot# RGT35619

	SAMPLE #	Type	Matrix	Initial Amount	Final Volume	Initial Vessel Wt	Final Vessel Wt	Spike Amount	Due Date
1	WG573656-02	BLANK	1	40 mL	50 mL	206.478 g	206.475 g		
2	WG573656-03	LCS	1	40 mL	50 mL	213.551 g	213.556 g	5 mL	
3	L16060840-01	SAMP	1	40 mL	50 mL	205.261 g	205.248 g		06/29/16
4	L16060840-02	SAMP	1	40 mL	50 mL	205.378 g	205.365 g		06/29/16
5	L16060840-03	SAMP	1	40 mL	50 mL	206.649 g	206.637 g		06/29/16
6	L16060840-04	SAMP	1	40 mL	50 mL	205.952 g	205.94 g		06/29/16
7	L16060840-05	SAMP	1	40 mL	50 mL	204.164 g	204.152 g		06/29/16
8	L16060840-06	SAMP	1	40 mL	50 mL	204.89 g	204.875 g		06/29/16
9	WG573656-01	REF	1	40 mL	50 mL	205.243 g	205.233 g		
10	L16060996-01	RS01	1	40 mL	50 mL	205.243 g	205.233 g		07/01/16
11	WG573656-04	MS	1	40 mL	50 mL	209.524 g	209.507 g	5 mL	
12	L16060996-02	MS01	1	40 mL	50 mL	209.524 g	209.507 g	5 mL	07/01/16
13	WG573656-05	MSD	1	40 mL	50 mL	211.803 g	211.788 g	5 mL	
14	L16060996-03	SD01	1	40 mL	50 mL	211.803 g	211.788 g	5 mL	07/01/16
15	L16060996-04	SAMP	1	40 mL	50 mL	203.833 g	203.826 g		07/01/16
16	L16060996-05	SAMP	1	40 mL	50 mL	207.088 g	207.078 g		07/01/16
17	L16060996-06	SAMP	1	40 mL	50 mL	205.731 g	205.723 g		07/01/16
18	L16060996-07	SAMP	1	40 mL	50 mL	204.58 g	204.57 g		07/01/16
19	L16060996-08	SAMP	1	40 mL	50 mL	204.442 g	204.434 g		07/01/16
20	L16060996-09	SAMP	1	40 mL	50 mL	206.06 g	206.051 g		07/01/16
21	L16061026-02	SAMP	1	40 mL	50 mL	207.271 g	207.264 g		06/28/16
22	L16061026-04	SAMP	1	40 mL	50 mL	207.035 g	207.028 g		06/28/16
23	L16061026-06	SAMP	1	40 mL	50 mL	205.716 g	205.709 g		06/28/16
24	L16061026-07	SAMP	1	40 mL	50 mL	206.892 g	206.885 g		06/28/16
25	L16061116-23	SAMP	1	40 mL	50 mL	205.051 g	205.04 g		07/01/16
26	L16061116-24	SAMP	1	40 mL	50 mL	207.895 g	207.88 g		07/01/16
27	L16061179-01	SAMP	1	40 mL	50 mL	208.78 g	208.772 g		06/28/16

L16061026-06	filtered digestate
L16061026-07	filtered digestate

Analyst: Erin Potter

Reviewer: Vicki Collins

MW_DIG - Modified 09/30/2009
PDF ID: 4817590
Report generated: 06/22/2016 09:06

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062716T3.1R.TXT
Analyst1: JYH Analyst2: N/A
Method: 200.7/6010B/6010C SOP: ME600G Rev: 8
Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838
ICSA: STD76382 ICSAB: STD76834 Int. Std: RGT31668
CCV: STD76511 LLCCV: STD76601 Tuning Sol :
Stannous : Hydroxylamine :

Workgroups: 574214,574076

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	T3.062716.101109	WG574382-01	Calibration Point		1		06/27/16 10:11
2	T3.062716.101506	WG574382-02	Calibration Point		1		06/27/16 10:15
3	T3.062716.101903	WG574382-03	Calibration Point		1		06/27/16 10:19
4	T3.062716.102300	WG574382-04	Calibration Point		1		06/27/16 10:23
5	T3.062716.102637	WG574382-05	Calibration Point		1		06/27/16 10:26
6	T3.062716.103013	WG574382-06	Initial Calibration Verification		1		06/27/16 10:30
7	T3.062716.103339	WG574382-07	Initial Calib Blank		1		06/27/16 10:33
8	T3.062716.103737	WG574382-08	Low Level Initial Calibration V		1		06/27/16 10:37
9	T3.062716.104131	WG574382-09	Interference Check		1		06/27/16 10:41
10	T3.062716.104522	WG574382-10	Interference Check		1		06/27/16 10:45
11	T3.062716.104904	WG574382-11	CCV		1		06/27/16 10:49
12	T3.062716.105241	WG574382-12	CCB		1		06/27/16 10:52
13	T3.062716.105639	WG573656-02	Method/Prep Blank	40/50	1		06/27/16 10:56
14	T3.062716.110423	WG573656-03	Laboratory Control S	40/50	1		06/27/16 11:04
15	T3.062716.110802	WG573656-01	Reference Sample		1	L16060996-01	06/27/16 11:08
16	T3.062716.111157	WG573656-04	Matrix Spike	40/50	1	L16060996-01	06/27/16 11:11
17	T3.062716.111535	WG573656-05	Matrix Spike Duplica	40/50	1	L16060996-01	06/27/16 11:15
18	T3.062716.111914	L16061179-01	13273-G06-WQ-W0218	40/50	1		06/27/16 11:19
19	T3.062716.112312	WG574214-01	Post Digestion Spike		1	L16061179-01	06/27/16 11:23
20	T3.062716.112651	WG574214-02	Serial Dilution		5	L16061179-01	06/27/16 11:26
21	T3.062716.113048	WG574214-02	Serial Dilution		25	L16061179-01	06/27/16 11:30
22	T3.062716.113445	WG574382-13	CCV		1		06/27/16 11:34
23	T3.062716.113822	WG574382-14	CCB		1		06/27/16 11:38
24	T3.062716.114220	L16061026-02	18WW16F-061616	40/50	1		06/27/16 11:42
25	T3.062716.114614	L16061026-04	129-061616	40/50	1		06/27/16 11:46
26	T3.062716.115009	L16061026-06	120F-061616	40/50	1		06/27/16 11:50
27	T3.062716.115400	L16061026-07	123-061616	40/50	1		06/27/16 11:54
28	T3.062716.115753	L16060840-01	MW-24-004R	40/50	1		06/27/16 11:57
29	T3.062716.120148	L16060840-02	MW-24-003R	40/50	1		06/27/16 12:01
30	T3.062716.120539	L16060840-03	FR-061416-24	40/50	1		06/27/16 12:05
31	T3.062716.120930	L16060840-04	MW-24-002R	40/50	1		06/27/16 12:09
32	T3.062716.121323	L16060840-05	MW-24-005	40/50	1		06/27/16 12:13
33	T3.062716.121733	L16060840-06	EB-LOW FLOW TUBING	40/50	1		06/27/16 12:17
34	T3.062716.122130	WG574382-15	CCV		1		06/27/16 12:21

Page: 1 Approved: June 28, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-THERMO3 Dataset: 062716T3.1R.TXT

Analyst1: JYH Analyst2: N/A

Method: 200.7/6010B/6010C SOP: ME600G Rev: 8

Maintenance Log ID:

Calibration Std: STD76513 ICV Std: STD76512 Post Spike: STD75838

ICSA: STD76382 ICSAB: STD76834 Int. Std: RGT31668

CCV: STD76511 LLCCV: STD76601 Tuning Sol :

Stannous : Hydroxylamine :

Workgroups: 574214,574076

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	T3.062716.122507	WG574382-16	CCB		1		06/27/16 12:25
36	T3.062716.122905	L16060996-04	MW-21-006	40/50	1		06/27/16 12:29
37	T3.062716.123259	L16060996-05	MW-21-002	40/50	1		06/27/16 12:32
38	T3.062716.123653	L16060996-06	FD-061516-21	40/50	1		06/27/16 12:36
39	T3.062716.124047	L16060996-07	MW-21-009	40/50	1		06/27/16 12:40
40	T3.062716.124440	L16060996-08	MW-21-005	40/50	1		06/27/16 12:44
41	T3.062716.124834	L16060996-09	MW-21-004	40/50	1		06/27/16 12:48
42	T3.062716.125227	L16061116-23	062016-PS-CB027	40/50	1		06/27/16 12:52
43	T3.062716.125618	L16061116-24	062016-PS-CB027	40/50	1		06/27/16 12:56
44	T3.062716.130008	L16061026-02	18WW16F-061616	40/50	20		06/27/16 13:00
45	T3.062716.130358	L16061026-04	129-061616	40/50	20		06/27/16 13:03
46	T3.062716.130753	WG574382-17	CCV		1		06/27/16 13:07
47	T3.062716.131130	WG574382-18	CCB		1		06/27/16 13:11
48	T3.062716.131528	L16061026-06	120F-061616	40/50	20		06/27/16 13:15
49	T3.062716.131923	L16061290-03	MW16F-062116	40/50	20		06/27/16 13:19
50	T3.062716.132320	WG574382-19	CCV		1		06/27/16 13:23
51	T3.062716.132646	WG574382-20	CCB		1		06/27/16 13:26
52	T3.062716.133045	WG574382-21	Low Level Continuing Calibra		1		06/27/16 13:30

Page: 2 Approved: June 28, 2016

Kim H. Rhoden

Microbac

Microbac Laboratories Inc.

Data Checklist

Date: 27-JUN-2016
Analyst: JYH
Analyst: NA
Method: 6010B/6010C/200.7
Instrument: ICP-THERMO3
Curve Workgroup: 574382
Runlog ID: 75950
Analytical Workgroups: 574656,574076

Add'l WG	
STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	X
ICB/CCB	X
ICSA/ICSAB	X
CRI	X
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	X
Client Forms	X
Level X	
Level 3	840,996
Level 4	1026,1179,1290
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	JYH
Secondary Reviewer	KHR
Comments	

Primary Reviewer:

Secondary Reviewer:
28-JUN-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUN-28-2016 10:41:09

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 6010C

AAB#: WG574214

Login Number: L16061026

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18WW16F-061616	02	06/16/16					06/22/2016	6	180		06/27/16	11.2	180	
18WW16F-061616	02	06/16/16					06/22/2016	6	180		06/27/16	11.2	180	
129-061616	04	06/16/16					06/22/2016	6	180		06/27/16	11.2	180	
129-061616	04	06/16/16					06/22/2016	6	180		06/27/16	11.1	180	
120F-061616	06	06/16/16					06/22/2016	5.9	180		06/27/16	11.2	180	
120F-061616	06	06/16/16					06/22/2016	5.9	180		06/27/16	11.1	180	
123-061616	07	06/16/16					06/22/2016	5.9	180		06/27/16	11.1	180	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4827991
Report generated 06/28/2016 09:31

METHOD BLANK SUMMARY

Login Number: L16061026
Blank File ID: T3.062716.105639
Prep Date: 06/22/16 07:53
Analyzed Date: 06/27/16 10:56
Analyst: JYH

Work Group: WG574214
Blank Sample ID: WG573656-02
Instrument ID: ICP-THERMO3
Method: 6010C

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573656-03	T3.062716.110423	06/27/16 11:04	01
18WW16F-061616	L16061026-02	T3.062716.114220	06/27/16 11:42	01
129-061616	L16061026-04	T3.062716.114614	06/27/16 11:46	01
120F-061616	L16061026-06	T3.062716.115009	06/27/16 11:50	01
123-061616	L16061026-07	T3.062716.115400	06/27/16 11:54	01
18WW16F-061616	L16061026-02	T3.062716.130008	06/27/16 13:00	DL01
129-061616	L16061026-04	T3.062716.130358	06/27/16 13:03	DL01
120F-061616	L16061026-06	T3.062716.131528	06/27/16 13:15	DL01

Report Name: BLANK_SUMMARY
PDF File ID: 4827992
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16061026 Prep Date: 06/22/16 07:53 Sample ID: WG573656-02
Instrument ID: ICP-THERMO3 Run Date: 06/27/16 10:56 Prep Method: 3015
File ID: T3.062716.105639 Analyst: JYH Method: 6010C
Workgroup (AAB#): WG574214 Matrix: Water Units: mg/L
Contract #: Cal ID: ICP-TH-27-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Aluminum, Total	0.0500	0.200	0.0500	1	U
Beryllium, Total	0.00500	0.0200	0.00500	1	U
Calcium, Total	0.125	0.500	0.125	1	U
Iron, Total	0.0500	0.200	0.0500	1	U
Magnesium, Total	0.250	1.00	0.250	1	U
Potassium, Total	0.500	2.00	0.500	1	U
Selenium, Total	0.00500	0.0200	-0.00914	1	U
Sodium, Total	0.250	1.00	0.250	1	U

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4827993
28-JUN-2016 09:31

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16061026 Run Date: 06/27/2016 Sample ID: WG573656-03
Instrument ID: ICP-THERMO3 Run Time: 11:04 Prep Method: 3015
File ID: T3.062716.110423 Analyst: JYH Method: 6010C
Workgroup (AAB#): WG574214 Matrix: Water Units: mg/L
QC Key: DOD4 Lot#: STD76576 Cal ID: ICP-TH-27-JUN-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Aluminum, Total	6.25	5.90	94.4	80 - 120	
Beryllium, Total	0.0313	0.0292	93.4	80 - 120	
Calcium, Total	6.25	6.02	96.4	80 - 120	
Iron, Total	2.50	2.39	95.4	80 - 120	
Magnesium, Total	6.25	6.02	96.3	80 - 120	
Potassium, Total	31.3	29.8	95.4	80 - 120	
Selenium, Total	0.250	0.235	93.8	80 - 120	
Sodium, Total	31.3	30.2	96.6	80 - 120	

LCS - Modified 03/06/2008
PDF File ID: 4827994
Report generated: 06/28/2016 09:31

Microbac Laboratories Inc.
MATRIX SPIKE AND MATRIX SPIKE DUP (MS/MSD)

Loginnum: L16061026 Cal ID: ICP-THERMO3 - Worknum: WG574214
Instrument ID: ICP-THERMO3 Contract #: Method: 6010C
Parent ID: WG573656-01 File ID: T3.062716.110802 Dil: 1 Matrix: WATER
Sample ID: WG573656-04 MS File ID: T3.062716.111157 Dil: 1 Units: mg/L
Sample ID: WG573656-05 MSD File ID: T3.062716.111535 Dil: 1

Analyte	Parent	MS Spiked	MS Found	MS %Rec	MSD Spiked	MSD Found	MSD %Rec	%RPD	%Rec Limits	RPD Limit	Q
Aluminum	ND	6.25	5.92	94.7	6.25	5.98	95.7	1.09	80 - 120	20	
Beryllium	ND	0.0313	0.0293	93.6	0.0313	0.0293	93.8	0.213	80 - 120	20	
Calcium	70.2	6.25	76.5	100	6.25	76.1	95.1	0.411	80 - 120	20	
Iron	ND	2.50	2.41	96.3	2.50	2.41	96.6	0.301	80 - 120	20	
Magnesium	28.5	6.25	34.1	89.6	6.25	34.1	89.2	0.0733	80 - 120	20	
Potassium	ND	31.3	30.2	96.6	31.3	30.2	96.7	0.0579	80 - 120	20	
Selenium	ND	0.250	0.228	91.3	0.250	0.236	94.3	3.28	80 - 120	20	
Sodium	5.16	31.3	35.4	96.6	31.3	35.3	96.4	0.212	80 - 120	20	

* FAILS %REC LIMIT

FAILS RPD LIMIT

NOTE: This is an internal quality control sample.

WG_MS_MSD_DRYWT - Modified 05/26/2011
PDF File ID: 4827995
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
Serial Dilution Report

Login: L16061026 **Worknum:** WG574214
Instrument: ICP-THERMO3 **Method:** 6010C
Serial Dil: WG574214-02 **File ID:** T3.062716.112651 **Dil:** 5 **Units:** ug/L
Sample: L16061179-01 **File ID:** T3.062716.111914 **Dil:** 1

Analyte	Sample	Qual	Serial Dil	Qual	% Diff	Q
Aluminum	20.8		ND	U		
Beryllium	ND	U	ND	U		
Calcium	186		98.0		47.20	E
Iron	29.3		101		243.00	E
Magnesium	74.1		ND	U		
Potassium	74.5		887		1090.00	E
Selenium	ND	U	ND	U		
Sodium	151		545		261.00	E

U = Result is below MDL.
F = Result is greater than or equal to MDL and less than the RL.
X = Result is greater than or equal to RL and less than 25 times the MDL.
E = %D exceeds control limit of 10% and initial sample result is greater than or equal to 25 times the MDL.

SERIAL_DIL - Modified 09/22/2008
PDF File ID: 4827989
06/28/2016 09:31

Microbac Laboratories Inc.
POST SPIKE REPORT

Sample Login ID: L16061026 **Worknum:** WG574214
Instrument ID: ICP-THERMO3 **Method:** 6010C
Post Spike ID: WG574214-01 **File ID:** T3.062716.112312 **Dil:** 1 **Units:** ug/L
Sample ID: L16061179-01 **File ID:** T3.062716.111914 **Dil:** 1 **Matrix:** Water

Analyte	Post Spike Result	C	Sample Result	C	Spike Added(SA)	% R	Control Limit %R	Q
ALUMINUM	4690		0	U	5000	93.9	75 - 125	
BERYLLIUM	23.0		0	U	25	92.1	75 - 125	
CALCIUM	4980		186	F	5000	96.2	75 - 125	
IRON	1910		0	U	2000	95.6	75 - 125	
MAGNESIUM	4720		0	U	5000	94.4	75 - 125	
POTASSIUM	23400		0	U	25000	93.7	75 - 125	
SELENIUM	179		0	U	200	89.7	75 - 125	
SODIUM	23800		0	U	25000	95.4	75 - 125	

N = % Recovery exceeds control limits
F = Result is between MDL and RL
U = Sample result is below MDL. A value of zero is used in the calculation

POST_SPIKE - Modified 03/06/2008
PDF File ID: 4827990
Report generated: 06/28/2016 09:31

Microbac Laboratories Inc.
Initial Calibration Summary

Login:	L16061026	Workgroup (AAB#):	WG574214
Analytical Method:	6010C	Instrument ID:	ICP-THERMO3
ICAL Worknum:	WG574382	Initial Calibration Date:	27-JUN-2016 10:26

	WG574382-01		WG574382-02		WG574382-03		WG574382-04		WG574382-05			
	Conc	INT	Conc	INT	Conc	INT	Conc	INT	Conc	INT	R	Q
ALUMINUM	0	0.000800	.1	0.00149	.2	0.00217	10	0.0861	20	0.161	.999541	
BERYLLIUM	0	0.000690	.0005	0.00111	.001	0.00148	.05	0.0508	.1	0.0959	.999682	
CALCIUM	0	0.00169	.1	0.00256	.2	0.00981	10	0.474	20	0.899	.998908	
IRON	0	0.0000700	.04	0.000210	.08	0.00147	4	0.0959	8	0.181	.998751	
MAGNESIUM	0	-0.0000400	NA	NA	.2	0.00164	10	0.0916	20	0.172	.999436	
POTASSIUM	0	-0.0166	.5	0.00590	1	0.0309	50	2.79	100	5.30	.999643	
SELENIUM	0	-0.000420	NA	NA	.008	-0.000340	.4	0.00633	.8	0.0123	.999365	
SODIUM	0	-0.0629	.5	0.0187	1	0.0937	50	9.16	100	17.5	.999663	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
*** = Out of Compliance; R < 0.995**

INT_CAL_ICP - Modified 03/06/2008
PDF File ID: 4827998
Report generated: 28-JUN-2016 09:31

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-07
Instrument ID: ICP-THERMO3
Run Time: 10:33
Method: 6010C
File ID: T3.062716.103339
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG574214
Cal ID: ICP-THERI - 27-JUN-16
Matrix: WATER

Analytes	MDL	RDL	Concentration	Qualifier
ALUMINUM	.04	.16	.04	U
BERYLLIUM	.004	.016	.004	U
CALCIUM	.1	.4	.1	U
IRON	.04	.16	.04	U
MAGNESIUM	.2	.8	.2	U
POTASSIUM	.4	1.6	.4	U
SELENIUM	.004	.016	-.00421	F
SODIUM	.2	.8	.2	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4828000
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-12
Instrument ID: ICP-THERMO3
Run Time: 10:52
Method: 6010C
File ID: T3.062716.105241
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	-0.00430	F
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828003
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-14
Instrument ID: ICP-THERMO3
Run Time: 11:38
Method: 6010C
File ID: T3.062716.113822
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828003
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-16
Instrument ID: ICP-THERMO3
Run Time: 12:25
Method: 6010C
File ID: T3.062716.122507
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828003
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-18
Instrument ID: ICP-THERMO3
Run Time: 13:11
Method: 6010C
File ID: T3.062716.131130
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828003
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-20
Instrument ID: ICP-THERMO3
Run Time: 13:26
Method: 6010C
File ID: T3.062716.132646
Analyst: JYH
Units: mg/L
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Aluminum	0.0400	0.160	0.0400	U
Beryllium	0.00400	0.0160	0.00400	U
Calcium	0.100	0.400	0.100	U
Iron	0.0400	0.160	0.0400	U
Magnesium	0.200	0.800	0.200	U
Potassium	0.400	1.60	0.400	U
Selenium	0.00400	0.0160	0.00400	U
Sodium	0.200	0.800	0.200	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4828003
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16061026 Run Date: 06/27/2016 Sample ID: WG574382-06
Instrument ID: ICP-THERMO3 Run Time: 10:30 Method: 6010C
File ID: T3.062716.103013 Analyst: JYH Units: mg/L
Workgroup (AAB#): WG574214 Cal ID: ICP-TH - 27-JUN-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Aluminum		10	10.0	100	90 - 110	
Beryllium		.05	0.0495	99.0	90 - 110	
Calcium		10	9.89	98.9	90 - 110	
Iron		4	3.95	98.7	90 - 110	
Magnesium		10	10.0	100	90 - 110	
Potassium		50	49.5	99.0	90 - 110	
Selenium		.4	0.405	101	90 - 110	
Sodium		50	49.6	99.1	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4827999
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-11
Instrument ID: ICP-THERMO3
Run Time: 10:49
Method: 6010C
File ID: T3.062716.104904
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	10.7	mg/L	107	90 - 110		
Beryllium		0.0500	0.0516	mg/L	103	90 - 110		
Calcium		10.0	10.8	mg/L	108	90 - 110		
Iron		4.00	4.30	mg/L	108	90 - 110		
Magnesium		10.0	10.8	mg/L	108	90 - 110		
Potassium		50.0	51.5	mg/L	103	90 - 110		
Selenium		0.400	0.421	mg/L	105	90 - 110		
Sodium		50.0	51.6	mg/L	103	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828002
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-13
Instrument ID: ICP-THERMO3
Run Time: 11:34
Method: 6010C
File ID: T3.062716.113445
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	10.3	mg/L	103	90 - 110		
Beryllium		0.0500	0.0508	mg/L	102	90 - 110		
Calcium		10.0	10.2	mg/L	102	90 - 110		
Iron		4.00	4.09	mg/L	102	90 - 110		
Magnesium		10.0	10.3	mg/L	103	90 - 110		
Potassium		50.0	50.8	mg/L	102	90 - 110		
Selenium		0.400	0.410	mg/L	103	90 - 110		
Sodium		50.0	51.2	mg/L	102	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828002
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-15
Instrument ID: ICP-THERMO3
Run Time: 12:21
Method: 6010C
File ID: T3.062716.122130
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	10.3	mg/L	103	90 - 110		
Beryllium		0.0500	0.0508	mg/L	102	90 - 110		
Calcium		10.0	10.9	mg/L	109	90 - 110		
Iron		4.00	4.34	mg/L	109	90 - 110		
Magnesium		10.0	10.8	mg/L	108	90 - 110		
Potassium		50.0	53.5	mg/L	107	90 - 110		
Selenium		0.400	0.416	mg/L	104	90 - 110		
Sodium		50.0	54.6	mg/L	109	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828002
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-17
Instrument ID: ICP-THERMO3
Run Time: 13:07
Method: 6010C
File ID: T3.062716.130753
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	9.98	mg/L	99.8	90 - 110		
Beryllium		0.0500	0.0490	mg/L	97.9	90 - 110		
Calcium		10.0	9.54	mg/L	95.4	90 - 110		
Iron		4.00	3.77	mg/L	94.1	90 - 110		
Magnesium		10.0	9.47	mg/L	94.7	90 - 110		
Potassium		50.0	47.1	mg/L	94.2	90 - 110		
Selenium		0.400	0.421	mg/L	105	90 - 110		
Sodium		50.0	48.1	mg/L	96.2	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828002
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-19
Instrument ID: ICP-THERMO3
Run Time: 13:23
Method: 6010C
File ID: T3.062716.132320
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		10.0	9.90	mg/L	99.0	90 - 110		
Beryllium		0.0500	0.0485	mg/L	97.0	90 - 110		
Calcium		10.0	9.79	mg/L	97.9	90 - 110		
Iron		4.00	3.90	mg/L	97.4	90 - 110		
Magnesium		10.0	9.50	mg/L	95.0	90 - 110		
Potassium		50.0	47.8	mg/L	95.6	90 - 110		
Selenium		0.400	0.409	mg/L	102	90 - 110		
Sodium		50.0	49.3	mg/L	98.7	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4828002
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-08
Instrument ID: ICP-THERMO3
Run Time: 10:37
Method: 6010C
File ID: T3.062716.103737
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		0.160	0.161	mg/L	101	70 - 130		
Beryllium		0.00160	0.00171	mg/L	107	70 - 130		
Calcium		0.400	0.413	mg/L	103	70 - 130		
Iron		0.0800	0.0618	mg/L	77.3	70 - 130		
Magnesium		0.400	0.370	mg/L	92.5	70 - 130		
Potassium		0.800	0.831	mg/L	104	70 - 130		
Selenium		0.0160	0.0156	mg/L	97.2	70 - 130		
Sodium		0.400	0.456	mg/L	114	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4828004
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16061026
Run Date: 06/27/2016
Sample ID: WG574382-21
Instrument ID: ICP-THERMO3
Run Time: 13:30
Method: 6010C
File ID: T3.062716.133045
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574214
Cal ID: ICP-TH - 27-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Aluminum		0.160	0.168	mg/L	105	70 - 130		
Beryllium		0.00160	0.00157	mg/L	98.1	70 - 130		
Calcium		0.400	0.410	mg/L	102	70 - 130		
Iron		0.0800	0.0608	mg/L	76.0	70 - 130		
Magnesium		0.400	0.392	mg/L	98.0	70 - 130		
Potassium		0.800	0.912	mg/L	114	70 - 130		
Selenium		0.0160	0.0148	mg/L	92.2	70 - 130		
Sodium		0.400	0.499	mg/L	125	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4828004
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16061026
Instrument ID: ICP-THERMO3
Sol. A : WG574382-09 **File ID:** T3.062716.104131
Sol. AB : WG574382-10 **File ID:** T3.062716.104522

Workgroup (AAB#): WG574214
Method: 6010C
Units: mg/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Aluminum	250	274	110	250	256	102	
Beryllium	NS	-0.0000500	NS	0.250	0.237	94.8	
Calcium	250	255	102	250	239	95.6	
Iron	100	101	101	100	93.5	93.5	
Magnesium	250	258	103	250	239	95.6	
Potassium	NS	-0.0902	NS	5.00	5.12	102	
Selenium	NS	-0.00510	NS	0.250	0.245	98.0	
Sodium	NS	0.0834	NS	5.00	5.19	104	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4828001
Report generated 06/28/2016 09:31

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16061026 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	AG	AL	AS	B	BA
ALUMINUM	308.20	0	0	0	0	0
ANTIMONY	206.80	0	0.0000410	0	0	0
ARSENIC	189.00	0	0	0	0	0
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	0	0	0	0
BORON	249.60	0	0	0	0	0
CADMIUM	228.80	0	0	0.0115	0	-0.0000800
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0	0	0	0	0
COBALT	228.60	0	0	0	0	0
COPPER	224.70	0	0	0	0	0
IRON	261.10	0	0	0	0	0
LEAD	220.30	0	0.000260	0	0	0
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	0	0	0	0	0
MANGANESE	257.60	0	0	0	0	0
MOLYBDENUM	202.00	0	0	0	0	0
NICKEL	231.60	0	0	0	0	0
PHOSPHORUS	214.90	0	-0.000289	0	0	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0	-0.0000490	0	0	0
SILICON	212.40	0	0	0	0	0
SILVER	328.00	0	0	0	0	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0	0	0	0
THALLIUM	190.80	0	-0.0000120	0	0	0
TIN	189.90	0	0	0	0	0
TITANIUM	337.20	0	0	0	0	0
VANADIUM	292.40	0	0	0	0	0
ZINC	206.20	0	0.0000300	0	0	0
ZIRCONIUM	339.10	0	0	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4827997
Report generated: 06/28/2016 09:31

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16061026 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	BE	CA	CD	CO	CR
ALUMINUM	308.20	0	0	0	-0.000820	0
ANTIMONY	206.80	0	0	0	0	0.0200
ARSENIC	189.00	0	0	0	0	-0.00190
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	0	0	0	0
BORON	249.60	0	0	0	0.00343	0
CADMIUM	228.80	0	0	0	-0.00390	0
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0	0	0	0	0
COBALT	228.60	0	0	0	0	-0.000200
COPPER	224.70	0	0	0	0.0000770	-0.00100
IRON	261.10	0	0	0	0	-0.00100
LEAD	220.30	0	0	0	-0.0000130	-0.000132
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	0	0	0	0	0
MANGANESE	257.60	0	0	0	0	0.0000500
MOLYBDENUM	202.00	0	0	0	0	0
NICKEL	231.60	0	0	0	-0.000860	0
PHOSPHORUS	214.90	0	0	0	0	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0	0	0	0	0
SILICON	212.40	0	0	0	0	0
SILVER	328.00	0	0	0	0	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0.00000500	0	0	0
THALLIUM	190.80	0	0	0	0.00240	0.000276
TIN	189.90	0	0	0	0	0
TITANIUM	337.20	0	0	0	0	0
VANADIUM	292.40	0	0	0	0	-0.00480
ZINC	206.20	0	0	0	0	-0.00180
ZIRCONIUM	339.10	0	0	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4827997
Report generated: 06/28/2016 09:31

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16061026 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	CU	FE	K	LI	MG
ALUMINUM	308.20	0	0	0	0	0
ANTIMONY	206.80	0	0.0000560	0	0	0
ARSENIC	189.00	0	-0.0000500	0	0	0
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	0	0	0	0
BORON	249.60	0	0.000300	0	0	0
CADMIUM	228.80	0	-0.0000190	0	0	0
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0	0.0000500	0	0	0
COBALT	228.60	0	0	0	0	0
COPPER	224.70	0	0.00160	0	0	0
IRON	261.10	0	0	0	0	0
LEAD	220.30	0.000609	0	0	0	0
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	0	0	0	0	0
MANGANESE	257.60	0	0	0	0	0.00000300
MOLYBDENUM	202.00	0	0	0	0	0
NICKEL	231.60	0	0.0000420	0	0	0
PHOSPHORUS	214.90	-0.323	0.000900	0	0	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0	0	0	0	0
SILICON	212.40	0	0	0	0	0
SILVER	328.00	0	-0.000270	0	0	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0	0	0	0
THALLIUM	190.80	0	0	0	0	0
TIN	189.90	0	0	0	0	0
TITANIUM	337.20	0	-0.000400	0	0	0
VANADIUM	292.40	0	0.00000700	0	0	0
ZINC	206.20	0	0	0	0	0
ZIRCONIUM	339.10	0	-0.0000300	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4827997
Report generated: 06/28/2016 09:31

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16061026 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	MN	MO	NA	NI	P
ALUMINUM	308.20	0	0.0163	0	0	0
ANTIMONY	206.80	0	-0.00310	0	-0.00350	0
ARSENIC	189.00	0	0.00120	0	0	0
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	0	0	0	0
BORON	249.60	0	-0.00190	0	0	0
CADMIUM	228.80	0	0.0000320	0	-0.000770	0
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0.000360	0	0	0	0
COBALT	228.60	0	-0.00200	0	0.000100	0
COPPER	224.70	0	0.00160	0	-0.0123	0
IRON	261.10	0	0	0	0	0
LEAD	220.30	0	-0.00210	0	0.000110	0
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	-0.00290	-0.0230	0	0	0
MANGANESE	257.60	0	0.0000300	0	0	0
MOLYBDENUM	202.00	0	0	0	0	0
NICKEL	231.60	0	0	0	0	0
PHOSPHORUS	214.90	0	0.00710	0	0	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0.000600	0.000580	0	0	0
SILICON	212.40	0	0.0187	0	0	0
SILVER	328.00	0	-0.0000430	0	0	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0	0	0	0
THALLIUM	190.80	0.00100	0	0	0	0
TIN	189.90	0	0	0	0	0
TITANIUM	337.20	0	-0.000153	0	0	0
VANADIUM	292.40	-0.000200	-0.00830	0	0	0
ZINC	206.20	0	0	0	0	0
ZIRCONIUM	339.10	0	0	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4827997
Report generated: 06/28/2016 09:31

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16061026 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	PB	SB	SE	SI	SN
ALUMINUM	308.20	0	0	0	0	0
ANTIMONY	206.80	0	0	0	0	-0.0220
ARSENIC	189.00	0	0	0	0	0
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	0	0	0	0
BORON	249.60	0	0	0	0	0
CADMIUM	228.80	0	0	0	0	0
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0	0	0	0	0
COBALT	228.60	0	0	0	0	0
COPPER	224.70	0.00440	0	0	0	0
IRON	261.10	0	0	0	0	0
LEAD	220.30	0	0	0	0	0
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	0	0	0	0	0
MANGANESE	257.60	0	0	0	0	0
MOLYBDENUM	202.00	0	0	0	0	0
NICKEL	231.60	0	0	0	0	0
PHOSPHORUS	214.90	0	0	0	0	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0	0	0	0	0
SILICON	212.40	0	0	0	0	0
SILVER	328.00	0	0	0	0	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0	0	0	0
THALLIUM	190.80	0	0	0	0	0
TIN	189.90	0	0	0	0	0
TITANIUM	337.20	0	0	0	0	0
VANADIUM	292.40	0	0	0	0	0
ZINC	206.20	0	0	0	0	0
ZIRCONIUM	339.10	0	0	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4827997
Report generated: 06/28/2016 09:31

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16061026 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	SR	TI	TL	V	ZN
ALUMINUM	308.20	0	0	0	0.0950	0
ANTIMONY	206.80	0	0.00110	0	-0.00360	0
ARSENIC	189.00	0	0	0	0.000107	0
BARIUM	455.40	0	0	0	0	0
BERYLLIUM	313.10	0	-0.00000700	0	0.000990	0
BORON	249.60	0	0	0	0	0
CADMIUM	228.80	0	0	0	0.000102	0
CALCIUM	422.60	0	0	0	0	0
CHROMIUM	267.70	0	0.0000550	0	0	0
COBALT	228.60	0	0.00210	0	0.0000200	0
COPPER	224.70	0	0.000269	0	0	0
IRON	261.10	0	0	0	0	0
LEAD	220.30	0	0	0	-0.000126	0
LITHIUM	670.70	0	0	0	0	0
MAGNESIUM	279.00	0	-0.00290	0	0	0
MANGANESE	257.60	0	0	0	0	0
MOLYBDENUM	202.00	0	0	0	-0.000110	0
NICKEL	231.60	0	0	0	0	0
PHOSPHORUS	214.90	0	0	0	-0.00100	0
POTASSIUM	766.40	0	0	0	0	0
SELENIUM	196.00	0	0	0	0	0
SILICON	212.40	0	0	0	0	0
SILVER	328.00	0	-0.000720	0	-0.000260	0
SODIUM	589.50	0	0	0	0	0
STRONTIUM	407.70	0	0	0	0	0
THALLIUM	190.80	0	-0.000800	0	-0.00490	0
TIN	189.90	0	-0.00190	0	0	0
TITANIUM	337.20	0	0	0	0	0
VANADIUM	292.40	0	0.000820	0	0	0
ZINC	206.20	0	0	0	0	0
ZIRCONIUM	339.10	0	0	0	0	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4827997
Report generated: 06/28/2016 09:31

Microbac Laboratories Inc.
INTERELEMENT CORRECTION FACTORS (ANNUALLY)

Login Number: L16061026 **Date:** 01/04/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Wave Length	ZR
ALUMINUM	308.20	0
ANTIMONY	206.80	0
ARSENIC	189.00	0
BARIUM	455.40	0
BERYLLIUM	313.10	0
BORON	249.60	0
CADMIUM	228.80	0
CALCIUM	422.60	0
CHROMIUM	267.70	0
COBALT	228.60	0
COPPER	224.70	0
IRON	261.10	0
LEAD	220.30	0
LITHIUM	670.70	0
MAGNESIUM	279.00	0
MANGANESE	257.60	0
MOLYBDENUM	202.00	0
NICKEL	231.60	0
PHOSPHORUS	214.90	0
POTASSIUM	766.40	0
SELENIUM	196.00	0
SILICON	212.40	0
SILVER	328.00	0
SODIUM	589.50	0
STRONTIUM	407.70	0
THALLIUM	190.80	0
TIN	189.90	0
TITANIUM	337.20	0
VANADIUM	292.40	0
ZINC	206.20	0
ZIRCONIUM	339.10	0

CORR_FACTORS - Modified 03/05/2008
PDF File ID: 4827997
Report generated: 06/28/2016 09:31

Microbac Laboratories Inc.
LINEAR RANGE (QUARTERLY)

Login Number: L16061026 **Date:** 04/29/2016
Insturment ID: ICP-THERMO3 **Method:** 6010C

Analyte	Integration Time (Sec.)	Concentration (mg/L)
Aluminum	10.00	900.0
Antimony	20.00	45.0
Arsenic	10.00	45.0
Barium	10.00	45.0
Beryllium	10.00	1.8
Boron	20.00	45.0
Cadmium	20.00	4.5
Calcium	5.00	270.0
Chromium	20.00	36.0
Cobalt	20.00	45.0
Copper	20.00	180.0
Iron	5.00	720.0
Lead	20.00	225.0
Lithium	5.00	36.0
Magnesium	5.00	900.0
Manganese	10.00	36.0
Molybdenum	20.00	27.0
Nickel	20.00	90.0
Phosphorus	20.00	180.0
Potassium	5.00	450.0
Selenium	20.00	90.0
Silicon	20.00	36.0
Silver	10.00	9.0
Sodium	5.00	270.0
Strontium	5.00	9.0
Thallium	20.00	18.0
Tin	20.00	45.0
Titanium	5.00	36.0
Vanadium	20.00	27.0
Zinc	20.00	45.0
Zirconium	10.00	45.0

Comments:
All analytes passed acceptance criteria at the specified concentration.

LINEAR_RANGE - Modified 03/06/2008
PDF File ID: 4827996
Report generated: 06/28/2016 09:31

2.3.1.3 Raw Data

Element, Wavelength and Order	Date of Fit	Date of Cal.	Type of Fit	Weighting	A0	A1	A2	n (Exponent)
Ag 328.068 {103}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000219	0.051332	0.000000	1.000000
Al 308.215 {109}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000797	0.008110	0.000000	1.000000
As 189.042 {478}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000055	0.023360	0.000000	1.000000
B 249.678 {135}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000001	0.022663	0.000000	1.000000
Ba 455.403 { 74}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.055586	2.146038	0.000000	1.000000
Be 313.107 {108}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000688	0.950408	0.000000	1.000000
Ca 422.673 { 80}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.001692	0.045598	0.000000	1.000000
Cd 228.802 {447}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000094	0.563612	0.000000	1.000000
Co 228.616 {447}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000104	0.359063	0.000000	1.000000
Cr 267.716 {126}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000207	0.060248	0.000000	1.000000
Cu 224.700 {450}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000063	0.126227	0.000000	1.000000
Fe 261.187 {129}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000073	0.023032	0.000000	1.000000
K 766.490 { 44}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.016555	0.054237	0.000000	1.000000
Li 670.784 { 50}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.004397	1.544241	0.000000	1.000000
Mg 279.079 {121}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000044	0.008813	0.000000	1.000000
Mn 257.610 {131}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000249	0.281817	0.000000	1.000000
Mo 202.030 {467}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000046	0.182222	0.000000	1.000000
Na 589.592 { 57}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.062909	0.178536	0.000000	1.000000
Ni 231.604 {446}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000706	0.133269	0.000000	1.000000
P 214.914 {457}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000179	0.013467	0.000000	1.000000
Pb 220.353 {453}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000979	0.062395	0.000000	1.000000
Sb 206.833 {463}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000288	0.029684	0.000000	1.000000
Se 196.090 {472}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000416	0.016174	0.000000	1.000000
Si 212.412 {459}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.002457	0.038614	0.000000	1.000000
Sn 189.989 {477}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000032	0.070434	0.000000	1.000000
Sr 407.771 { 83}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000984	3.576768	0.000000	1.000000
Ti 337.280 {100}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.001213	0.097488	0.000000	1.000000
Tl 190.856 {477}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000404	0.028714	0.000000	1.000000
V 292.402 {115}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.000025	0.070640	0.000000	1.000000
Y 224.306 {450}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Y 360.073 { 94}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Y 377.433 { 89}*	<not fit>	<Never Calibrated>	Linear	1/Conc	0.000000	0.000000	0.000000	1.000000
Zn 206.200 {463}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	0.000222	0.414339	0.000000	1.000000
Zr 339.198 { 99}	6/27/2016 11:02:26	6/27/2016 10:30:08	Linear	1/Conc	-0.004799	3.597030	0.000000	1.000000

Approved: June 28, 2016

Element, Wavelength and Order	Correlation	Std Error of Est	Predicted MDL	Predicted MQL	Status	Reslope		QC Norm	
						Slope	Y-int	Slope factor	Offset
Ag 328.068 {103}	0.999219	0.000005	0.002404	0.008013	OK.	1.000000	0.000000	1	0
Al 308.215 {109}	0.999541	0.000016	0.010486	0.034953	OK.	1.000000	0.000000	1	0
As 189.042 {478}	0.999717	0.000002	0.003514	0.011712	OK.	1.000000	0.000000	1	0
B 249.678 {135}	0.999559	0.000003	0.003229	0.010763	OK.	1.000000	0.000000	1	0
Ba 455.403 { 74}	0.999622	0.000374	0.001188	0.003959	OK.	1.000000	0.000000	1	0
Be 313.107 {108}	0.999682	0.000008	0.000105	0.000351	OK.	1.000000	0.000000	1	0
Ca 422.673 { 80}	0.998908	0.000135	0.046468	0.154893	OK.	1.000000	0.000000	1	0
Cd 228.802 {447}	0.999582	0.000006	0.000295	0.000984	OK.	1.000000	0.000000	1	0
Co 228.616 {447}	0.999651	0.000012	0.000445	0.001484	OK.	1.000000	0.000000	1	0
Cr 267.716 {126}	0.999210	0.000008	0.001312	0.004375	OK.	1.000000	0.000000	1	0
Cu 224.700 {450}	0.999350	0.000015	0.001489	0.004962	OK.	1.000000	0.000000	1	0
Fe 261.187 {129}	0.998751	0.000029	0.029643	0.098810	OK.	1.000000	0.000000	1	0
K 766.490 { 44}	0.999643	0.000460	0.132623	0.442075	OK.	1.000000	0.000000	1	0
Li 670.784 { 50}	0.999776	0.000321	0.004353	0.014511	OK.	1.000000	0.000000	1	0
Mg 279.079 {121}	0.999436	0.000029	0.087066	0.290220	OK.	1.000000	0.000000	1	0
Mn 257.610 {131}	0.999089	0.000038	0.002851	0.009505	OK.	1.000000	0.000000	1	0
Mo 202.030 {467}	0.999789	0.000024	0.000466	0.001553	OK.	1.000000	0.000000	1	0
Na 589.592 { 57}	0.999663	0.001472	0.035366	0.117888	OK.	1.000000	0.000000	1	0
Ni 231.604 {446}	0.999563	0.000013	0.001271	0.004238	OK.	1.000000	0.000000	1	0
P 214.914 {457}	0.999580	0.000024	0.007845	0.026150	OK.	1.000000	0.000000	1	0
Pb 220.353 {453}	0.998800	0.000010	0.003957	0.013189	OK.	1.000000	0.000000	1	0
Sb 206.833 {463}	0.999499	0.000007	0.004891	0.016304	OK.	1.000000	0.000000	1	0
Se 196.090 {472}	0.999365	0.000002	0.007423	0.024744	OK.	1.000000	0.000000	1	0
Si 212.412 {459}	0.999891	0.000018	0.002581	0.008604	OK.	1.000000	0.000000	1	0
Sn 189.989 {477}	0.999675	0.000011	0.000958	0.003192	OK.	1.000000	0.000000	1	0
Sr 407.771 { 83}	0.999546	0.000684	0.000493	0.001644	OK.	1.000000	0.000000	1	0
Ti 337.280 {100}	0.997531	0.000043	0.009062	0.030207	OK.	1.000000	0.000000	1	0
Tl 190.856 {477}	0.999388	0.000005	0.003390	0.011298	OK.	1.000000	0.000000	1	0
V 292.402 {115}	0.999660	0.000012	0.001079	0.003598	OK.	1.000000	0.000000	1	0
Y 224.306 {450}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Y 360.073 { 94}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Y 377.433 { 89}*	0.000000	0.000000	-1.000000	-1.000000	Warnin	1.000000	0.000000	1	0
Zn 206.200 {463}	0.999413	0.000090	0.000220	0.000735	OK.	1.000000	0.000000	1	0
Zr 339.198 { 99}	0.999407	0.000786	0.000564	0.001879	OK.	1.000000	0.000000	1	0

Approved: June 28, 2016

Sample Name: S0 Acquired: 6/27/2016 10:11:09 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00022**	**.00080**	**.00005**	**-.00000**	**.05558**	**.00069**	**.00169**
Stddev	.00009	.00005	.00001	.00001	.00343	.00010	.00171
%RSD	41.038	6.2901	19.225	757.31	6.1761	14.508	100.89
#1	-.00028	.00085	.00004	.00001	.05677	.00057	.00305
#2	-.00012	.00079	.00007	-.00001	.05827	.00074	-.00023
#3	-.00027	.00075	.00006	.00000	.05171	.00075	.00226

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00009**	**-.00010**	**.00021**	**.00006**	**.00007**	**-.01656**	**-.00440**
Stddev	.00009	.00008	.00005	.00012	.00021	.00039	.00634
%RSD	91.046	78.768	26.095	189.05	283.56	2.3394	144.17
#1	-.00005	-.00014	.00014	.00010	.00028	-.01656	.00211
#2	-.00019	-.00016	.00023	.00016	.00009	-.01694	-.00474
#3	-.00004	-.00001	.00024	-.00007	-.00014	-.01617	-.01056

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00004**	**.00025**	**-.00005**	**-.06293**	**-.00071**	**-.00018**	**-.00098**
Stddev	.00014	.00078	.00008	.01073	.00010	.00001	.00026
%RSD	309.31	312.25	167.27	17.052	14.737	6.2600	26.253
#1	-.00014	.00113	-.00005	-.06678	-.00074	-.00017	-.00122
#2	.00011	-.00035	.00003	-.05080	-.00079	-.00019	-.00101
#3	-.00010	-.00004	-.00012	-.07120	-.00059	-.00018	-.00071

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00029**	**-.00042**	**.00246**	**.00003**	**.00098**	**-.00121**	**-.00040**
Stddev	.00006	.00008	.00007	.00002	.00168	.00048	.00011
%RSD	21.252	20.303	2.7310	73.870	172.37	39.424	26.113
#1	.00028	-.00050	.00242	.00001	.00288	-.00094	-.00028
#2	.00035	-.00033	.00242	.00005	-.00031	-.00176	-.00045
#3	.00023	-.00042	.00253	.00003	.00036	-.00093	-.00047

Approved: June 28, 2016

Sample Name: S0 Acquired: 6/27/2016 10:11:09 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S
Avg	**-.00002**	**.00022**	**-.00480**
Stddev	.00006	.00004	.00081
%RSD	232.69	19.281	16.811
#1	.00001	.00027	-.00516
#2	.00000	.00020	-.00387
#3	-.00009	.00019	-.00536
Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9360.7**	**56289.**	**2789.5**
Stddev	12.0	199.	9.9
%RSD	.12862	.35359	.35524
#1	9347.7	56429.	2778.2
#2	9362.8	56061.	2796.7
#3	9371.5	56376.	2793.6

Approved: June 28, 2016

Sample Name: S1 Acquired: 6/27/2016 10:15:06 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	Ba4554	Be3131	Ca4226	Cd2288	Co2286
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00002**	**.00149**	**.07501**	**.00111**	**.00256**	**.00019**	**.00048**
Stddev	.00004	.00003	.00130	.00003	.00054	.00000	.00012
%RSD	235.36	1.8291	1.7274	3.0313	21.065	2.0861	24.410
#1	.00003	.00146	.07531	.00115	.00264	.00019	.00036
#2	-.00003	.00151	.07359	.00111	.00199	.00019	.00059
#3	.00005	.00151	.07613	.00108	.00306	.00020	.00049

Elem	Cr2677	Cu2247	Fe2611	K_7664	Mn2576	Mo2020	Na5895
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00052**	**.00073**	**.00021**	**.00590**	**.00180**	**.00157**	**.01873**
Stddev	.00010	.00013	.00039	.00261	.00100	.00005	.00232
%RSD	18.633	17.713	184.86	44.235	55.870	3.0472	12.407
#1	.00062	.00086	.00004	.00630	.00167	.00152	.02125
#2	.00042	.00060	.00065	.00312	.00086	.00158	.01668
#3	.00052	.00074	-.00007	.00829	.00286	.00161	.01826

Elem	Ni2316	P_2149	Pb2203	Sb2068	Si2124	Sn1899	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00019**	**.00098**	**-.00091**	**.00069**	**.00415**	**.00064**	**.03417**
Stddev	.00010	.00003	.00038	.00011	.00003	.00003	.00071
%RSD	53.461	3.4756	41.630	15.291	.77728	4.6599	2.0672
#1	-.00029	.00101	-.00109	.00069	.00416	.00068	.03441
#2	-.00009	.00099	-.00047	.00080	.00418	.00063	.03337
#3	-.00018	.00095	-.00116	.00059	.00412	.00062	.03472

Elem	Ti3372	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00142**	**.00056**	**.00379**	**.02139**
Stddev	.00038	.00006	.00002	.00234
%RSD	26.838	9.9829	.42124	10.947
#1	-.00118	.00061	.00377	.02369
#2	-.00121	.00058	.00380	.02146
#3	-.00186	.00050	.00379	.01901

Approved: June 28, 2016

Sample Name: S1 Acquired: 6/27/2016 10:15:06 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9406.4**	**56869.**	**2813.2**
Stddev	17.9	99.	5.4
%RSD	.18981	.17489	.19258
#1	9415.6	56895.	2818.5
#2	9417.7	56760.	2813.3
#3	9385.8	56954.	2807.7

Approved: June 28, 2016

Sample Name: S2 Acquired: 6/27/2016 10:19:03 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.00019	.00217	.00019	.00017	.09097	.00148	.00981
Stddev	.00010	.00007	.00004	.00006	.00192	.00005	.00183
%RSD	52.118	3.1223	20.733	37.133	2.1077	3.6422	18.643
#1	.00022	.00221	.00022	.00021	.09288	.00154	.01182
#2	.00008	.00209	.00021	.00019	.09098	.00149	.00936
#3	.00027	.00220	.00015	.00010	.08905	.00143	.00824

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.00046	.00101	.00085	.00125	.00147	.03094	.02040
Stddev	.00011	.00020	.00003	.00016	.00011	.00596	.00709
%RSD	24.286	20.196	3.9657	12.594	7.8265	19.261	34.737
#1	.00054	.00105	.00082	.00140	.00135	.03776	.02303
#2	.00050	.00119	.00084	.00127	.00157	.02830	.01238
#3	.00033	.00079	.00088	.00108	.00148	.02676	.02580

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.00164	.00196	.00309	.09374	.00041	.00222	-.00046
Stddev	.00038	.00079	.00004	.01209	.00013	.00011	.00029
%RSD	23.225	40.222	1.2531	12.896	30.782	4.8080	63.126
#1	.00162	.00249	.00314	.08010	.00046	.00212	-.00025
#2	.00127	.00233	.00308	.09800	.00050	.00233	-.00079
#3	.00203	.00105	.00307	.10312	.00026	.00221	-.00034

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.00091	-.00034	.00573	.00119	.06491	.00017	-.00015
Stddev	.00010	.00006	.00002	.00005	.00081	.00072	.00004
%RSD	10.635	17.225	.29680	4.1503	1.2526	428.70	27.714
#1	.00083	-.00029	.00572	.00116	.06568	.00089	-.00012
#2	.00102	-.00033	.00575	.00125	.06406	-.00056	-.00013
#3	.00090	-.00040	.00572	.00117	.06500	.00018	-.00019

Approved: June 28, 2016

Sample Name: S2 Acquired: 6/27/2016 10:19:03 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S
Avg	.00120	.00738	.04944
Stddev	.00007	.00006	.00070
%RSD	5.6773	.87197	1.4116
#1	.00118	.00744	.05008
#2	.00114	.00738	.04954
#3	.00127	.00732	.04870

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9443.2	56709.	2845.2
Stddev	3.5	174.	6.8
%RSD	.03751	.30616	.23990
#1	9441.6	56633.	2837.4
#2	9447.3	56587.	2849.5
#3	9440.8	56908.	2848.9

Approved: June 28, 2016

Sample Name: S3 Acquired: 6/27/2016 10:23:00 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.02092	.08605	.00965	.01177	2.2814	.05082	.47399	.03171
Stddev	.00009	.00009	.00006	.00007	.0081	.00011	.00141	.00016
%RSD	.44413	.10972	.62223	.59949	.35348	.21174	.29780	.50827
#1	.02097	.08616	.00961	.01185	2.2766	.05071	.47378	.03182
#2	.02098	.08602	.00972	.01174	2.2908	.05092	.47550	.03179
#3	.02081	.08598	.00962	.01172	2.2769	.05085	.47271	.03153

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.07428	.03147	.06632	.09593	2.7929	1.5861	.09155	.14801
Stddev	.00025	.00009	.00009	.00042	.0165	.0024	.00035	.00046
%RSD	.34137	.28140	.13410	.43596	.58971	.15475	.37815	.31281
#1	.07440	.03141	.06640	.09607	2.7817	1.5848	.09185	.14757
#2	.07444	.03143	.06635	.09545	2.8118	1.5890	.09164	.14849
#3	.07398	.03157	.06622	.09625	2.7852	1.5847	.09117	.14798

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.18692	9.1586	.06869	.13735	.03188	.03618	.00633	.19982
Stddev	.00095	.0246	.00038	.00018	.00005	.00008	.00006	.00030
%RSD	.50865	.26851	.55910	.13312	.14375	.21004	.95433	.15185
#1	.18754	9.1579	.06868	.13721	.03187	.03616	.00627	.19957
#2	.18740	9.1836	.06909	.13756	.03193	.03626	.00639	.20015
#3	.18583	9.1344	.06832	.13729	.03184	.03611	.00632	.19973

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.07280	3.7149	.09996	.01331	.07292	.43380	3.7667
Stddev	.00018	.0060	.00049	.00007	.00019	.00078	.0189
%RSD	.24836	.16167	.49462	.49011	.25771	.17890	.50287
#1	.07265	3.7109	.09960	.01337	.07280	.43456	3.7451
#2	.07300	3.7218	.09976	.01324	.07314	.43381	3.7806
#3	.07274	3.7121	.10052	.01332	.07283	.43301	3.7744

Approved: June 28, 2016

Sample Name: S3 Acquired: 6/27/2016 10:23:00 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9721.9**	**58099.**	**2930.8**
Stddev	18.9	74.	10.9
%RSD	.19470	.12678	.37352
#1	9743.1	58020.	2942.5
#2	9706.5	58111.	2920.7
#3	9716.1	58166.	2929.2

Approved: June 28, 2016

Sample Name: S4 Acquired: 6/27/2016 10:26:37 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.03974	.16089	.01833	.02213	4.2645	.09586	.89929	.06038
Stddev	.00005	.00019	.00006	.00006	.0077	.00012	.00503	.00018
%RSD	.11654	.11956	.34341	.25502	.18096	.12203	.55878	.29985
#1	.03968	.16079	.01837	.02214	4.2575	.09600	.89526	.06018
#2	.03976	.16111	.01826	.02207	4.2633	.09579	.89769	.06053
#3	.03977	.16076	.01837	.02218	4.2728	.09580	.90492	.06045

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.14063	.05916	.12449	.18125	5.3047	3.0377	.17177	.27547
Stddev	.00052	.00006	.00037	.00104	.0273	.0084	.00079	.00147
%RSD	.37329	.09570	.29429	.57556	.51409	.27796	.45809	.53422
#1	.14021	.05914	.12420	.18098	5.2909	3.0280	.17115	.27390
#2	.14121	.05922	.12491	.18037	5.2872	3.0420	.17150	.27568
#3	.14045	.05912	.12437	.18241	5.3361	3.0431	.17265	.27682

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.35921	17.472	.12976	.25968	.06042	.06915	.01227	.38616
Stddev	.00077	.044	.00041	.00013	.00014	.00023	.00002	.00037
%RSD	.21360	.25231	.31950	.04942	.23319	.32893	.13923	.09489
#1	.35987	17.454	.12931	.25960	.06034	.06889	.01229	.38588
#2	.35938	17.440	.13013	.25983	.06033	.06932	.01226	.38658
#3	.35837	17.522	.12985	.25962	.06058	.06923	.01226	.38603

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.13808	7.0066	.19048	.02529	.13828	.80821	7.0194
Stddev	.00024	.0219	.00266	.00011	.00020	.00129	.0190
%RSD	.17717	.31279	1.3974	.43280	.14820	.16017	.27028
#1	.13814	6.9980	.18870	.02521	.13848	.80908	7.0103
#2	.13829	6.9903	.18921	.02525	.13830	.80882	7.0066
#3	.13781	7.0315	.19354	.02541	.13807	.80672	7.0412

Approved: June 28, 2016

Sample Name: S4 Acquired: 6/27/2016 10:26:37 Type: Cal
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: IR Corr. Factor: 1.000000
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9771.3**	**58546.**	**2991.1**
Stddev	5.3	169.	6.3
%RSD	.05411	.28803	.21196
#1	9765.6	58357.	2985.0
#2	9776.0	58602.	2997.6
#3	9772.3	58680.	2990.8

Approved: June 28, 2016

Sample Name: ICV Acquired: 6/27/2016 10:30:13 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39656	9.9968	.39605	.49913	.99354	.04949	9.8900	.04960
Stddev	.00038	.0168	.00361	.00130	.00797	.00012	.1192	.00034
%RSD	.09463	.16774	.91201	.26037	.80254	.24834	1.2050	.67893
#1	.39683	10.016	.39921	.49920	.99329	.04957	9.8437	.04996
#2	.39672	9.9871	.39211	.49780	1.0016	.04955	10.025	.04956
#3	.39613	9.9872	.39682	.50040	.98569	.04935	9.8010	.04929
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.19960	.49616	.50076	3.9491	49.503	.98788	10.028	.49846
Stddev	.00072	.00107	.00153	.0174	.377	.01095	.058	.00163
%RSD	.36296	.21613	.30576	.43969	.76139	1.1083	.57412	.32766
#1	.20035	.49503	.50215	3.9291	49.521	.98444	9.9746	.49786
#2	.19954	.49627	.49912	3.9607	49.871	1.0001	10.089	.50030
#3	.19890	.49717	.50101	3.9574	49.117	.97907	10.020	.49721
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.98929	49.553	.50008	9.9383	.49711	1.1842	.40491	4.9048
Stddev	.00695	.361	.00310	.0420	.00341	.0012	.00380	.0152
%RSD	.70230	.72857	.61963	.42238	.68619	.10088	.93725	.30954
#1	.99664	49.590	.50317	9.9837	.49981	1.1844	.40809	4.9217
#2	.98842	49.894	.50009	9.9303	.49824	1.1853	.40593	4.9005
#3	.98282	49.175	.49697	9.9009	.49328	1.1829	.40071	4.8922
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Approved: June 28, 2016

Sample Name: ICV Acquired: 6/27/2016 10:30:13 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.99382**	**.99081**	**.99014**	**.50327**	**.99007**	**1.0044**	**.99998**
Stddev	.00477	.00901	.00242	.00334	.00183	.0042	.00712
%RSD	.47986	.90979	.24468	.66313	.18511	.41413	.71179
#1	.99906	.99161	.99266	.50658	.98958	1.0088	1.0007
#2	.99266	.99939	.98992	.50331	.99209	1.0036	1.0067
#3	.98974	.98142	.98783	.49991	.98852	1.0006	.99254
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9944.5**	**59548.**	**3007.2**
Stddev	7.0	104.	19.1
%RSD	.07028	.17468	.63490
#1	9939.5	59668.	3013.8
#2	9941.5	59493.	2985.7
#3	9952.5	59483.	3022.1

Approved: June 28, 2016

Sample Name: ICB Acquired: 6/27/2016 10:33:39 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00174	.00806	-.00035	.00524	.00096	.00003	.04488
Stddev	.00255	.00339	.00155	.00316	.00054	.00006	.03416
%RSD	146.61	42.061	438.03	60.356	56.377	169.93	76.128
#1	.00413	.01096	-.00078	.00453	.00153	.00002	.04662
#2	-.00095	.00889	-.00165	.00869	.00045	-.00001	.07814
#3	.00204	.00433	.00137	.00249	.00091	.00010	.00988
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00026	.00021	.00264	.00138	-.01027	.11431	-.00187
Stddev	.00023	.00025	.00090	.00135	.02697	.10019	.00400
%RSD	90.076	117.41	34.187	98.005	262.67	87.648	214.31
#1	.00019	.00011	.00242	.00087	.02024	.22955	-.00367
#2	.00007	.00003	.00363	.00292	-.02010	.06552	-.00465
#3	.00052	.00050	.00186	.00036	-.03093	.04786	.00272
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.07897	.00101	.00510	.11885	-.00064	F .01147	-.00335
Stddev	.04472	.00084	.00024	.04616	.00013	.00517	.00352
%RSD	56.634	82.917	4.6972	38.837	20.205	45.049	105.12
#1	-.03785	.00103	.00483	.16190	-.00071	.01200	-.00401
#2	-.12658	.00016	.00517	.12455	-.00049	.01635	.00046
#3	-.07246	.00185	.00530	.07011	-.00073	.00606	-.00650
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						.01000	
Low Limit						-.01000	

Approved: June 28, 2016

Sample Name: ICB Acquired: 6/27/2016 10:33:39 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00383	-.00421	.01837	.00133	.00214	-.00619	.00159
Stddev	.00530	.00215	.00326	.00121	.00026	.00445	.00067
%RSD	138.24	51.159	17.738	90.647	12.024	71.914	42.141
#1	.00877	-.00203	.02032	.00237	.00238	-.00553	.00121
#2	-.00176	-.00633	.02019	.00162	.00218	-.01094	.00119
#3	.00449	-.00427	.01461	.00001	.00187	-.00211	.00236
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.00117	.00107	.00150
Stddev	.00040	.00023	.00032
%RSD	33.968	21.646	21.450
#1	.00144	.00132	.00117
#2	.00136	.00103	.00151
#3	.00071	.00086	.00182
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9485.8	56960.	2827.2
Stddev	22.7	62.	13.4
%RSD	.23928	.10801	.47261
#1	9462.0	57031.	2842.2
#2	9488.3	56934.	2822.9
#3	9507.2	56917.	2816.6

Approved: June 28, 2016

Sample Name: LLICV Acquired: 6/27/2016 10:37:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00811	.16107	.00546	.08342	.00820	.00171	.41320	.00104
Stddev	.00154	.00831	.00174	.00283	.00174	.00005	.05309	.00032
%RSD	18.985	5.1608	31.869	3.3900	21.229	3.2040	12.848	30.976
#1	.00877	.17043	.00450	.08566	.00976	.00165	.44800	.00137
#2	.00635	.15823	.00748	.08024	.00632	.00173	.35210	.00072
#3	.00921	.15456	.00442	.08437	.00853	.00175	.43951	.00104
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00385	.00622	.00454	.06184	.83144	.08347	.36989	.00868
Stddev	.00035	.00111	.00053	.00835	.03960	.00277	.07407	.00380
%RSD	9.1544	17.869	11.677	13.497	4.7632	3.3221	20.024	43.760
#1	.00354	.00567	.00393	.06716	.81243	.08203	.29122	.00933
#2	.00376	.00549	.00491	.06614	.87696	.08170	.38018	.00459
#3	.00423	.00750	.00478	.05222	.80492	.08666	.43828	.01210
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00933	.45563	.01629	.80974	.00826	.08707	.01555	.84594
Stddev	.00021	.01951	.00128	.00898	.00183	.00241	.00253	.00721
%RSD	2.2323	4.2816	7.8796	1.1092	22.131	2.7708	16.251	.85208
#1	.00915	.46087	.01543	.81937	.00648	.08622	.01843	.85233
#2	.00956	.43404	.01777	.80827	.01013	.08979	.01372	.84735
#3	.00929	.47199	.01568	.80159	.00817	.08519	.01449	.83813
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: LLICV Acquired: 6/27/2016 10:37:37 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41453**	**.04310**	**.02153**	**.15768**	**.00889**	**.01738**	**.04625**
Stddev	.00755	.00023	.00874	.00411	.00040	.00018	.00135
%RSD	1.8207	.53307	40.597	2.6083	4.5428	1.0465	2.9262
#1	.41925	.04320	.01815	.15896	.00867	.01736	.04544
#2	.41852	.04284	.03146	.16100	.00865	.01758	.04550
#3	.40583	.04326	.01499	.15308	.00936	.01721	.04781
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9608.8**	**58008.**	**2882.7**
Stddev	13.9	202.	11.3
%RSD	.14503	.34866	.39311
#1	9595.5	57793.	2870.0
#2	9607.5	58037.	2891.6
#3	9623.3	58195.	2886.6

Approved: June 28, 2016

Sample Name: ICSA Acquired: 6/27/2016 10:41:31 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00348**	**274.46**	**-.00360**	**-.01905**	**-.00118**	**-.00005**	**255.24**	**-.00047**
Stddev	.00080	.12	.00424	.00363	.00140	.00007	.64	.00022
%RSD	22.958	.04489	117.79	19.069	118.85	159.46	.24922	45.496
#1	-.00266	274.32	-.00283	-.01618	-.00225	-.00002	254.54	-.00071
#2	-.00425	274.55	-.00817	-.01785	-.00168	-.00013	255.78	-.00029
#3	-.00353	274.52	.00020	-.02314	.00040	.00001	255.40	-.00042
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00138**	**.00013**	**.00026**	**100.86**	**-.09019**	**.00712**	**258.19**	**-.00082**
Stddev	.00020	.00043	.00219	.24	.06234	.00564	.51	.00449
%RSD	14.369	328.27	851.02	.23590	69.125	79.115	.19632	546.94
#1	-.00121	.00037	.00206	100.59	-.14551	.00643	257.65	-.00388
#2	-.00133	-.00036	.00088	100.95	-.02263	.00187	258.26	-.00293
#3	-.00159	.00038	-.00218	101.04	-.10243	.01307	258.66	.00434
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00009**	**.08336**	**-.00068**	**.02055**	**.00062**	**-.00085**	**-.00510**	**-.01322**
Stddev	.00049	.00057	.00126	.01032	.00616	.00460	.00665	.00103
%RSD	550.79	.68258	185.29	50.215	988.12	543.87	130.28	7.7725
#1	-.00016	.08380	-.00213	.03143	-.00603	.00410	-.01192	-.01441
#2	.00065	.08356	.00000	.01933	.00179	-.00500	.00136	-.01264
#3	-.00023	.08271	.00009	.01090	.00611	-.00163	-.00475	-.01262
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: ICSA Acquired: 6/27/2016 10:41:31 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00088	.00072	.00255	-.00142	-.00297	-.00684	-.00350
Stddev	.00092	.00022	.00227	.00469	.00113	.00030	.00043
%RSD	104.29	30.744	88.699	331.16	38.033	4.3723	12.250
#1	.00194	.00091	.00433	-.00152	-.00422	-.00699	-.00322
#2	.00027	.00078	.00000	-.00606	-.00263	-.00650	-.00330
#3	.00043	.00048	.00332	.00333	-.00204	-.00704	-.00400
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9003.1	53741.	2830.1
Stddev	29.5	88.	13.2
%RSD	.32793	.16418	.46682
#1	8969.0	53843.	2836.5
#2	9019.7	53686.	2838.9
#3	9020.6	53695.	2814.9

Approved: June 28, 2016

Sample Name: ICSAB Acquired: 6/27/2016 10:45:22 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.49885	256.39	.25252	-.01918	.24309	.23676	239.36	.48503
Stddev	.00077	.65	.00502	.00271	.00045	.00050	.30	.00068
%RSD	.15367	.25216	1.9898	14.136	.18678	.21093	.12664	.14053
#1	.49899	255.73	.25330	-.01721	.24333	.23629	239.67	.48508
#2	.49802	256.43	.25711	-.01805	.24256	.23671	239.35	.48569
#3	.49953	257.02	.24715	-.02227	.24337	.23729	239.06	.48433
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.23429	.24086	.24530	93.507	5.1189	.00241	239.40	.23738
Stddev	.00111	.00266	.00136	.226	.0691	.00187	.35	.00351
%RSD	.47486	1.1032	.55335	.24199	1.3498	77.806	.14736	1.4794
#1	.23518	.23891	.24433	93.596	5.0482	.00227	239.45	.24113
#2	.23465	.24388	.24472	93.675	5.1223	.00060	239.72	.23418
#3	.23304	.23978	.24685	93.249	5.1863	.00434	239.02	.23682
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00053	5.1902	.46938	.09211	.47966	.49223	.24503	-.02683
Stddev	.00007	.0094	.00026	.00814	.00226	.00746	.00251	.00281
%RSD	13.615	.18091	.05557	8.8363	.47050	1.5148	1.0231	10.477
#1	-.00046	5.1797	.46968	.08370	.48227	.49724	.24688	-.02680
#2	-.00060	5.1978	.46920	.09995	.47849	.49580	.24602	-.02404
#3	-.00052	5.1931	.46926	.09267	.47824	.48366	.24217	-.02966
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: ICSAB Acquired: 6/27/2016 10:45:22 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00063	.00073	-.00122	.45265	.24279	.46530	-.00351
Stddev	.00042	.00033	.00788	.00412	.00036	.00116	.00030
%RSD	67.531	45.343	645.29	.90940	.14968	.24968	8.4247
#1	.00101	.00064	-.00686	.45617	.24254	.46493	-.00322
#2	.00071	.00109	.00778	.44813	.24320	.46660	-.00349
#3	.00017	.00045	-.00458	.45366	.24261	.46437	-.00381
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9451.1	56687.	2980.7
Stddev	12.7	235.	6.1
%RSD	.13402	.41481	.20363
#1	9456.2	56956.	2985.6
#2	9436.6	56517.	2982.5
#3	9460.3	56590.	2973.9

Approved: June 28, 2016

Sample Name: CCV Acquired: 6/27/2016 10:49:04 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41238**	**10.669**	**.41191**	**.52005**	**1.0384**	**.05158**	**10.784**	**.05227**
Stddev	.00183	.046	.00331	.00113	.0055	.00021	.067	.00041
%RSD	.44263	.42702	.80236	.21781	.53089	.39861	.62379	.77509
#1	.41421	10.697	.41543	.52118	1.0324	.05173	10.736	.05274
#2	.41056	10.694	.40887	.52006	1.0433	.05167	10.860	.05201
#3	.41238	10.617	.41143	.51892	1.0395	.05134	10.754	.05206
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.20803**	**.51861**	**.51945**	**4.3037**	**51.532**	**1.0319**	**10.785**	**.52164**
Stddev	.00094	.00265	.00200	.0480	.292	.0030	.169	.00151
%RSD	.45047	.51067	.38573	1.1144	.56743	.29472	1.5692	.28971
#1	.20905	.51951	.52105	4.2715	51.286	1.0285	10.755	.52005
#2	.20720	.52068	.51721	4.3588	51.855	1.0344	10.967	.52182
#3	.20783	.51562	.52010	4.2807	51.456	1.0328	10.633	.52306
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.0260**	**51.605**	**.52017**	**10.355**	**.51948**	**1.2287**	**.42141**	**5.0896**
Stddev	.0062	.392	.00213	.026	.00211	.0063	.00578	.0141
%RSD	.60581	.75948	.40961	.24947	.40593	.51662	1.3721	.27654
#1	1.0327	51.153	.52203	10.384	.52081	1.2352	.42600	5.1052
#2	1.0247	51.808	.51785	10.341	.52057	1.2225	.42332	5.0779
#3	1.0205	51.853	.52065	10.339	.51705	1.2285	.41492	5.0858
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Approved: June 28, 2016

Sample Name: CCV Acquired: 6/27/2016 10:49:04 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.0329**	**1.0366**	**1.0326**	**.52078**	**1.0290**	**1.0446**	**1.0377**
Stddev	.0036	.0061	.0096	.00149	.0042	.0040	.0092
%RSD	.34765	.58644	.92987	.28568	.40577	.38739	.89149
#1	1.0370	1.0299	1.0217	.52198	1.0329	1.0493	1.0281
#2	1.0312	1.0418	1.0361	.52124	1.0295	1.0419	1.0386
#3	1.0304	1.0380	1.0399	.51911	1.0246	1.0427	1.0465
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9679.9**	**58019.**	**2940.7**
Stddev	31.4	99.	31.3
%RSD	.32391	.17011	1.0639
#1	9650.4	58117.	2973.3
#2	9676.5	57919.	2937.9
#3	9712.8	58020.	2910.9

Approved: June 28, 2016

Sample Name: CCB Acquired: 6/27/2016 10:52:41 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00082	.03709	-.00042	.00235	.00221	.00011	.02800	.00010
Stddev	.00285	.00947	.00333	.00050	.00064	.00007	.06022	.00006
%RSD	347.74	25.544	801.20	21.057	29.249	66.828	215.11	60.876
#1	.00282	.04765	-.00314	.00195	.00194	.00016	.00724	.00009
#2	.00208	.02934	.00330	.00220	.00174	.00014	.09585	.00005
#3	-.00244	.03428	-.00140	.00291	.00294	.00003	-.01910	.00017
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00002	.00256	-.00018	.01123	.17852	-.00287	.01155	.00102
Stddev	.00023	.00130	.00111	.02322	.06304	.00388	.06528	.00239
%RSD	1385.9	50.913	614.81	206.72	35.311	135.30	565.34	235.33
#1	-.00006	.00365	-.00083	-.00086	.24920	-.00008	.05161	.00372
#2	.00023	.00291	.00111	.03801	.15825	-.00730	.04682	.00017
#3	-.00022	.00112	-.00082	-.00345	.12812	-.00122	-.06378	-.00084
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00391	.11054	-.00006	.00936	-.00106	.00780	-.00430	.01094
Stddev	.00026	.02224	.00106	.00259	.00450	.00116	.00727	.00205
%RSD	6.5484	20.121	1773.4	27.704	423.55	14.865	169.26	18.686
#1	.00386	.10773	-.00045	.01132	-.00217	.00693	-.00997	.01182
#2	.00368	.13405	.00114	.01035	.00389	.00735	.00390	.00861
#3	.00419	.08984	-.00087	.00642	-.00490	.00911	-.00683	.01240
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: CCB Acquired: 6/27/2016 10:52:41 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00074**	**.00171**	**-.00592**	**.00236**	**.00119**	**.00045**	**.00147**
Stddev	.00073	.00018	.00639	.00105	.00056	.00012	.00032
%RSD	97.462	10.324	107.89	44.349	47.225	27.229	21.897
#1	.00158	.00178	-.00302	.00218	.00104	.00058	.00181
#2	.00028	.00184	-.01324	.00141	.00182	.00042	.00143
#3	.00037	.00151	-.00150	.00348	.00072	.00034	.00117
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9210.0**	**55620.**	**2799.3**
Stddev	17.9	95.	5.4
%RSD	.19389	.17128	.19191
#1	9214.9	55547.	2802.8
#2	9190.2	55728.	2793.2
#3	9224.9	55585.	2802.1

Approved: June 28, 2016

Sample Name: PBW AD Acquired: 6/27/2016 10:56:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00014	-.00551	-.00164	.00133	-.00255	.00001	-.02140	.00006
Stddev	.00131	.00261	.00242	.00099	.00017	.00008	.01807	.00008
%RSD	918.94	47.461	147.34	73.999	6.7565	679.86	84.478	126.18
#1	.00002	-.00708	-.00377	.00068	-.00272	-.00003	-.02206	-.00001
#2	-.00152	-.00249	.00099	.00085	-.00237	.00011	-.00300	.00014
#3	.00108	-.00696	-.00214	.00247	-.00257	-.00004	-.03913	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00013	.00044	-.00011	-.01481	.08157	-.00450	-.05709	-.00127
Stddev	.00038	.00059	.00174	.01969	.08524	.00309	.05809	.00048
%RSD	282.02	134.36	1518.6	132.89	104.49	68.616	101.75	38.034
#1	-.00023	.00035	.00020	.00330	.03640	-.00806	-.03041	-.00111
#2	.00028	-.00010	.00145	-.03577	.17989	-.00289	-.01713	-.00088
#3	-.00045	.00107	-.00199	-.01198	.02843	-.00255	-.12373	-.00181
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00120	.12315	-.00007	.00271	.00049	-.00196	-.00731	-.02994
Stddev	.00005	.02063	.00058	.00658	.00069	.00072	.00330	.00107
%RSD	4.3907	16.754	814.35	242.50	141.11	36.905	45.123	3.5596
#1	.00125	.13009	.00059	.01030	.00122	-.00130	-.00403	-.02923
#2	.00120	.13943	-.00033	-.00131	.00041	-.00184	-.01063	-.02942
#3	.00114	.09994	-.00048	-.00086	-.00016	-.00274	-.00728	-.03117
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Approved: June 28, 2016

Sample Name: PBW AD Acquired: 6/27/2016 10:56:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	.00122	.00010	.00267	.00012	.00086	.00027
Stddev	.00021	.00039	.00337	.00318	.00091	.00006	.00046
%RSD	418.85	32.122	3386.6	119.01	765.24	7.0777	171.03
#1	.00014	.00104	-.00194	.00621	.00001	.00079	.00037
#2	-.00019	.00095	.00399	.00178	.00108	.00091	.00068
#3	.00020	.00167	-.00175	.00003	-.00073	.00089	-.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10078.	61154.	3066.5
Stddev	38.	65.	16.3
%RSD	.37265	.10633	.53272
#1	10035.	61205.	3085.2
#2	10105.	61081.	3054.8
#3	10094.	61176.	3059.6

Approved: June 28, 2016

Sample Name: PBW AD Acquired: 6/27/2016 10:56:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00014**	**-.00551**	**-.00164**	**.00133**	**-.00255**	**.00001**	**-.02140**	**.00006**
Stddev	.00131	.00261	.00242	.00099	.00017	.00008	.01807	.00008
%RSD	907.10	47.461	147.33	74.064	6.7565	679.87	84.478	126.16
#1	.00002	-.00708	-.00377	.00068	-.00272	-.00003	-.02206	-.00001
#2	-.00153	-.00249	.00099	.00085	-.00237	.00011	-.00300	.00014
#3	.00108	-.00696	-.00214	.00247	-.00257	-.00004	-.03913	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00013**	**.00044**	**-.00011**	**-.01686**	**.08157**	**-.00450**	**-.05709**	**-.00127**
Stddev	.00038	.00059	.00176	.02442	.08524	.00309	.05809	.00048
%RSD	282.05	134.26	1612.9	144.86	104.49	68.616	101.75	38.034
#1	-.00023	.00035	.00021	-.00014	.03640	-.00806	-.03041	-.00111
#2	.00028	-.00010	.00147	-.04488	.17989	-.00289	-.01713	-.00088
#3	-.00045	.00107	-.00200	-.00555	.02843	-.00255	-.12373	-.00181
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00120**	**.12315**	**-.00007**	**.00272**	**.00049**	**-.00196**	**-.00731**	**-.02994**
Stddev	.00005	.02063	.00058	.00658	.00069	.00072	.00330	.00107
%RSD	4.3907	16.754	815.45	242.20	141.11	36.924	45.121	3.5596
#1	.00125	.13009	.00059	.01031	.00122	-.00130	-.00403	-.02923
#2	.00120	.13943	-.00033	-.00129	.00041	-.00184	-.01062	-.02942
#3	.00114	.09994	-.00048	-.00087	-.00016	-.00274	-.00727	-.03117
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: PBW AD Acquired: 6/27/2016 10:56:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v976) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00005	.00122	.00009	.00267	.00012	.00086	.00027
Stddev	.00021	.00039	.00336	.00318	.00091	.00006	.00046
%RSD	419.00	32.122	3582.0	119.01	765.15	7.0777	170.99
#1	.00014	.00104	-.00194	.00621	.00001	.00079	.00037
#2	-.00019	.00095	.00397	.00178	.00108	.00091	.00068
#3	.00020	.00167	-.00175	.00003	-.00073	.00089	-.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10078.	61154.	3066.5
Stddev	38.	65.	16.3
%RSD	.37265	.10633	.53272
#1	10035.	61205.	3085.2
#2	10105.	61081.	3054.8
#3	10094.	61176.	3059.6

Approved: June 28, 2016

Sample Name: LCSW AD Acquired: 6/27/2016 11:04:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-03

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.18930**	**4.7185**	**.19003**	**.93469**	**.47985**	**.02336**	**4.8179**	**.02393**
Stddev	.00264	.0066	.00083	.00065	.00055	.00013	.0255	.00027
%RSD	1.3934	.14053	.43752	.06932	.11364	.56886	.52848	1.1239
#1	.18653	4.7163	.19007	.93542	.48039	.02321	4.8423	.02405
#2	.19179	4.7132	.19083	.93448	.47985	.02342	4.7915	.02363
#3	.18958	4.7259	.18917	.93418	.47930	.02346	4.8198	.02412
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09659**	**.23846**	**.24088**	**1.9086**	**23.854**	**.48243**	**4.8127**	**.24088**
Stddev	.00021	.00006	.00176	.0299	.024	.00111	.0520	.00346
%RSD	.21848	.02654	.72986	1.5671	.09958	.22942	1.0798	1.4373
#1	.09683	.23854	.24210	1.8741	23.834	.48253	4.8448	.24488
#2	.09652	.23842	.24168	1.9261	23.880	.48128	4.7527	.23902
#3	.09642	.23843	.23887	1.9256	23.848	.48349	4.8404	.23875
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48317**	**24.148**	**.24158**	**4.6577**	**.24353**	**.56968**	**.18766**	**2.2805**
Stddev	.00185	.055	.00273	.0177	.00404	.00132	.00193	.0062
%RSD	.38336	.22645	1.1294	.38107	1.6575	.23248	1.0271	.27261
#1	.48319	24.206	.24282	4.6549	.24677	.56938	.18565	2.2798
#2	.48502	24.139	.24347	4.6766	.24481	.56853	.18785	2.2870
#3	.48131	24.098	.23845	4.6414	.23901	.57113	.18949	2.2746
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: LCSW AD Acquired: 6/27/2016 11:04:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-03

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.47831	.48449	.47945	.24451	.47571	.47996	.00474
Stddev	.00220	.00031	.00246	.00024	.00157	.00168	.00104
%RSD	.46005	.06434	.51251	.09744	.33103	.35074	21.927
#1	.47940	.48464	.47744	.24424	.47543	.48089	.00375
#2	.47976	.48413	.48219	.24467	.47429	.48098	.00463
#3	.47578	.48470	.47872	.24462	.47740	.47802	.00582
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10036.	60667.	3061.3
Stddev	16.	132.	7.4
%RSD	.15906	.21753	.24216
#1	10038.	60818.	3054.0
#2	10019.	60573.	3061.1
#3	10050.	60609.	3068.8

Approved: June 28, 2016

Sample Name: L1606099601 Acquired: 6/27/2016 11:08:02 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00124	.00326	-.00134	.00556	.00464	-.00007	56.162	.00040
Stddev	.00178	.00987	.00131	.00195	.00069	.00003	.529	.00030
%RSD	144.06	302.41	98.032	35.136	14.856	45.396	.94107	74.669
#1	.00220	.00071	.00016	.00782	.00529	-.00003	55.776	.00007
#2	.00234	.01416	-.00191	.00447	.00392	-.00009	55.945	.00048
#3	-.00082	-.00508	-.00226	.00439	.00472	-.00008	56.764	.00066
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00030	.00191	.00098	.00705	.36876	-.00116	22.804	.00093
Stddev	.00019	.00044	.00158	.01150	.11856	.00128	.343	.00214
%RSD	64.737	23.037	160.94	163.08	32.150	110.47	1.5021	228.81
#1	-.00024	.00235	.00067	.01641	.45961	-.00058	22.504	.00325
#2	-.00052	.00191	.00270	-.00578	.41204	-.00262	23.177	-.00096
#3	-.00015	.00147	-.00042	.01053	.23465	-.00027	22.730	.00051
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00111	4.1289	.00028	.00948	.00046	.00145	.00097	2.9475
Stddev	.00031	.0524	.00074	.00210	.00339	.00155	.00585	.0217
%RSD	27.978	1.2680	265.78	22.177	735.63	107.03	604.94	.73695
#1	.00110	4.0702	-.00001	.00869	.00432	.00119	-.00275	2.9626
#2	.00081	4.1708	.00111	.00789	-.00092	.00310	.00771	2.9573
#3	.00143	4.1456	-.00027	.01186	-.00202	.00004	-.00206	2.9226
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606099601 Acquired: 6/27/2016 11:08:02 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00048	.08076	-.01280	.00056	.00130	.00146	-.00019
Stddev	.00029	.00049	.00307	.00274	.00086	.00015	.00072
%RSD	60.177	.60538	24.011	488.83	65.984	10.053	371.62
#1	.00039	.08032	-.01635	-.00058	.00183	.00138	-.00101
#2	.00024	.08128	-.01096	.00369	.00175	.00163	.00038
#3	.00080	.08066	-.01109	-.00142	.00031	.00137	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10059.	60827.	3058.7
Stddev	29.	228.	2.0
%RSD	.29168	.37466	.06551
#1	10025.	60569.	3056.7
#2	10081.	61002.	3058.5
#3	10070.	60909.	3060.7

Approved: June 28, 2016

Sample Name: L1606099602S Acquired: 6/27/2016 11:11:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-04

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.18813**	**4.7350**	**.18948**	**.94279**	**.48703**	**.02341**	**61.170**	**.02398**
Stddev	.00277	.0076	.00172	.00254	.00188	.00011	.164	.00007
%RSD	1.4748	.16161	.90580	.26991	.38499	.47028	.26817	.30979
#1	.18980	4.7388	.19100	.94494	.48912	.02350	61.264	.02400
#2	.18493	4.7400	.18981	.93998	.48647	.02346	61.266	.02403
#3	.18966	4.7262	.18762	.94345	.48550	.02329	60.981	.02389
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.09458**	**.23494**	**.23535**	**1.9256**	**24.153**	**.48381**	**27.282**	**.23869**
Stddev	.00063	.00040	.00037	.0171	.111	.00423	.196	.00329
%RSD	.66218	.16824	.15837	.88843	.45831	.87512	.71904	1.3801
#1	.09454	.23528	.23512	1.9446	24.042	.48848	27.445	.23993
#2	.09522	.23450	.23515	1.9207	24.263	.48274	27.064	.24118
#3	.09397	.23504	.23578	1.9114	24.153	.48021	27.336	.23495
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.48631**	**28.289**	**.23480**	**4.7714**	**.23561**	**.57505**	**.18257**	**5.2951**
Stddev	.00099	.078	.00097	.0059	.00124	.00076	.00243	.0024
%RSD	.20421	.27704	.41396	.12448	.52582	.13199	1.3292	.04484
#1	.48717	28.373	.23464	4.7777	.23525	.57420	.18377	5.2952
#2	.48654	28.219	.23585	4.7707	.23699	.57531	.17978	5.2974
#3	.48523	28.274	.23393	4.7659	.23459	.57565	.18417	5.2926
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606099602S Acquired: 6/27/2016 11:11:57 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-04

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47604**	**.56589**	**.46385**	**.24049**	**.47449**	**.47286**	**.00122**
Stddev	.00004	.00229	.00771	.00260	.00102	.00053	.00044
%RSD	.00866	.40463	1.6627	1.0819	.21533	.11153	35.825
#1	.47602	.56845	.46505	.24323	.47561	.47256	.00081
#2	.47609	.56517	.45561	.23806	.47362	.47346	.00118
#3	.47602	.56404	.47090	.24019	.47423	.47254	.00168
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9956.7**	**60350.**	**3056.9**
Stddev	32.2	88.	8.3
%RSD	.32388	.14559	.27133
#1	9988.9	60290.	3047.9
#2	9924.4	60310.	3058.6
#3	9956.7	60451.	3064.2

Approved: June 28, 2016

Sample Name: L1606099603SD Acquired: 6/27/2016 11:15:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-06

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.19028	4.7868	.19269	.94867	.48722	.02346	60.919	.02406
Stddev	.00212	.0181	.00136	.00219	.00189	.00006	.253	.00014
%RSD	1.1126	.37744	.70818	.23039	.38720	.26793	.41487	.57233
#1	.19148	4.7722	.19359	.94673	.48507	.02353	60.675	.02401
#2	.18784	4.8070	.19112	.95104	.48802	.02344	60.902	.02395
#3	.19154	4.7811	.19336	.94824	.48858	.02341	61.180	.02421
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09518	.23692	.23715	1.9314	24.167	.48716	27.262	.23905
Stddev	.00047	.00079	.00143	.0184	.183	.00439	.044	.00192
%RSD	.49502	.33520	.60154	.95364	.75718	.90103	.16297	.80434
#1	.09469	.23782	.23571	1.9395	23.964	.48436	27.211	.23873
#2	.09522	.23634	.23856	1.9445	24.216	.48492	27.285	.23730
#3	.09563	.23660	.23717	1.9104	24.319	.49222	27.291	.24111
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.48886	28.229	.23606	4.7975	.24121	.57136	.18866	5.3330
Stddev	.00146	.184	.00176	.0017	.00169	.00269	.01260	.0141
%RSD	.29961	.65348	.74735	.03493	.70048	.47096	6.6800	.26466
#1	.49043	28.016	.23710	4.7973	.24241	.57255	.18833	5.3193
#2	.48863	28.338	.23706	4.7959	.24195	.57324	.20142	5.3475
#3	.48753	28.333	.23402	4.7993	.23928	.56827	.17622	5.3323
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606099603SD Acquired: 6/27/2016 11:15:35 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG573656-06

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47890**	**.56537**	**.47324**	**.24435**	**.47813**	**.47683**	**.00060**
Stddev	.00286	.00273	.01175	.00100	.00128	.00077	.00020
%RSD	.59645	.48334	2.4839	.40879	.26843	.16191	32.886
#1	.48155	.56309	.46094	.24471	.47955	.47751	.00072
#2	.47926	.56463	.47442	.24512	.47781	.47699	.00071
#3	.47587	.56840	.48436	.24322	.47704	.47599	.00037
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9937.3**	**60391.**	**3077.2**
Stddev	14.8	124.	21.3
%RSD	.14891	.20584	.69260
#1	9920.4	60281.	3101.8
#2	9943.6	60366.	3065.2
#3	9947.9	60526.	3064.6

Approved: June 28, 2016

Sample Name: L1606117901 Acquired: 6/27/2016 11:19:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00072	.02080	-.00015	.00912	-.00135	-.00001	.18554	.00009
Stddev	.00163	.01061	.00134	.00085	.00073	.00009	.03244	.00025
%RSD	227.50	50.989	907.80	9.2958	53.948	957.01	17.482	271.74
#1	.00097	.03252	.00133	.00831	-.00071	.00008	.19286	-.00017
#2	.00220	.01803	-.00129	.01000	-.00120	-.00010	.15007	.00033
#3	-.00102	.01186	-.00049	.00904	-.00214	-.00001	.21370	.00011
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00031	.00195	.00009	.02926	.07448	-.00083	.07412	.00218
Stddev	.00030	.00118	.00112	.04525	.16774	.00245	.02265	.00214
%RSD	96.395	60.268	1181.8	154.65	225.20	294.22	30.559	98.076
#1	-.00038	.00143	-.00002	-.01797	-.07824	-.00029	.04822	.00395
#2	.00002	.00330	.00126	.03352	.04769	.00130	.09021	.00278
#3	-.00058	.00112	-.00096	.07223	.25401	-.00351	.08393	-.00020
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00089	.15073	.00024	.00759	-.00104	.00154	-.00041	-.01205
Stddev	.00034	.03848	.00074	.00602	.00153	.00314	.00173	.00267
%RSD	38.727	25.532	313.77	79.298	146.52	204.07	419.45	22.167
#1	.00069	.14959	-.00056	.01447	.00072	.00196	.00111	-.00897
#2	.00069	.18977	.00089	.00499	-.00186	.00444	-.00229	-.01372
#3	.00129	.11282	.00038	.00331	-.00200	-.00179	-.00006	-.01346
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606117901 Acquired: 6/27/2016 11:19:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00000	.00123	-.00396	-.00083	.00060	.00107	-.00024
Stddev	.00073	.00029	.00070	.00214	.00012	.00021	.00050
%RSD	69743.	23.931	17.704	257.74	20.309	19.961	205.14
#1	.00084	.00157	-.00320	.00160	.00049	.00130	-.00015
#2	-.00035	.00106	-.00411	-.00168	.00073	.00102	-.00078
#3	-.00049	.00105	-.00458	-.00242	.00059	.00088	.00020
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10211.	62110.	3089.5
Stddev	13.	186.	5.8
%RSD	.12645	.29961	.18789
#1	10203.	62198.	3083.6
#2	10225.	61896.	3095.2
#3	10204.	62236.	3089.7

Approved: June 28, 2016

Sample Name: L1606117901PS Acquired: 6/27/2016 11:23:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG574214-01

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.18751	4.6924	.18741	.92180	.47420	.02302	4.9755	.02354
Stddev	.00138	.0184	.00060	.00040	.00158	.00005	.0105	.00025
%RSD	.73597	.39242	.32238	.04391	.33360	.22042	.21030	1.0799
#1	.18888	4.7119	.18682	.92179	.47602	.02303	4.9647	.02334
#2	.18612	4.6753	.18739	.92141	.47324	.02296	4.9856	.02383
#3	.18752	4.6899	.18803	.92222	.47333	.02306	4.9762	.02345
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.09588	.23472	.23888	1.9126	23.415	.47894	4.7182	.23827
Stddev	.00035	.00100	.00116	.0071	.048	.00290	.0903	.00057
%RSD	.36630	.42739	.48450	.37295	.20447	.60570	1.9132	.24122
#1	.09598	.23575	.23947	1.9074	23.375	.48172	4.6611	.23837
#2	.09616	.23375	.23962	1.9096	23.402	.47593	4.6712	.23766
#3	.09549	.23467	.23755	1.9207	23.468	.47917	4.8222	.23879
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.47661	23.842	.23804	4.6066	.23860	.55603	.17930	2.2773
Stddev	.00152	.022	.00175	.0111	.00481	.00366	.00848	.0068
%RSD	.31904	.09402	.73670	.24161	2.0142	.65792	4.7280	.29846
#1	.47744	23.817	.23835	4.6121	.23447	.55198	.16952	2.2796
#2	.47754	23.847	.23962	4.6139	.24387	.55702	.18392	2.2826
#3	.47486	23.861	.23616	4.5938	.23746	.55910	.18447	2.2696
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606117901PS Acquired: 6/27/2016 11:23:12 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment: WG574214-01

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.47165**	**.47956**	**.47085**	**.24029**	**.46670**	**.47457**	**.00055**
Stddev	.00140	.00029	.00956	.00263	.00189	.00170	.00039
%RSD	.29768	.06000	2.0311	1.0939	.40453	.35895	71.804
#1	.47125	.47964	.47754	.23905	.46790	.47446	.00081
#2	.47321	.47924	.47511	.23851	.46452	.47633	.00075
#3	.47048	.47980	.45990	.24331	.46767	.47292	.00010
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10135.**	**61356.**	**3086.5**
Stddev	4.	169.	6.2
%RSD	.04127	.27468	.20041
#1	10138.	61169.	3079.4
#2	10130.	61496.	3090.6
#3	10136.	61402.	3089.5

Approved: June 28, 2016

Sample Name: L1606117901SDL Acquired: 6/27/2016 11:26:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG574214-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00182	-.00876	-.00158	.00420	-.00145	-.00006	.01959	.00013
Stddev	.00139	.00831	.00054	.00197	.00159	.00007	.02852	.00015
%RSD	76.113	94.876	34.316	46.949	109.92	115.61	145.60	111.05
#1	.00305	.00028	-.00125	.00647	-.00089	.00000	.04998	.00030
#2	.00209	-.01048	-.00128	.00316	-.00021	-.00005	.01539	.00002
#3	.00032	-.01607	-.00220	.00296	-.00325	-.00013	-.00660	.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00024	.00078	.00077	.02010	.17744	-.00181	-.00514	.00019
Stddev	.00029	.00095	.00050	.00816	.07146	.00342	.03381	.00246
%RSD	124.09	121.80	65.402	40.597	40.275	189.49	657.34	1314.0
#1	-.00049	.00179	.00026	.01194	.25402	.00022	-.02869	.00024
#2	-.00029	-.00010	.00079	.02826	.11253	-.00576	-.02034	-.00230
#3	.00008	.00065	.00126	.02009	.16576	.00012	.03360	.00262
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00092	.10897	.00042	.00853	-.00135	.00022	-.00397	-.02369
Stddev	.00009	.00391	.00083	.00771	.00206	.00158	.00375	.00072
%RSD	9.7431	3.5859	195.85	90.342	152.42	719.67	94.434	3.0224
#1	.00096	.10458	.00127	.00472	.00097	.00168	-.00530	-.02420
#2	.00082	.11027	-.00039	.00348	-.00298	-.00146	.00026	-.02287
#3	.00099	.11207	.00039	.01741	-.00204	.00043	-.00687	-.02400
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606117901SDL Acquired: 6/27/2016 11:26:51 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 5 Custom ID2: Custom ID3:
Comment: WG574214-02

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00018	.00095	.00206	.00175	.00023	.00090	-.00011
Stddev	.00081	.00029	.00575	.00174	.00011	.00015	.00028
%RSD	447.31	30.155	279.47	99.665	49.179	16.523	251.35
#1	-.00057	.00062	.00863	.00026	.00024	.00096	-.00037
#2	.00104	.00111	-.00043	.00367	.00012	.00101	-.00014
#3	.00008	.00112	-.00203	.00132	.00035	.00073	.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10479.	63800.	3157.1
Stddev	55.	94.	17.6
%RSD	.52430	.14744	.55775
#1	10534.	63713.	3156.5
#2	10480.	63899.	3175.0
#3	10424.	63787.	3139.8

Approved: June 28, 2016

Sample Name: L1606117901SDL Acquired: 6/27/2016 11:30:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG574214-02

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00005**	**-.01463**	**-.00147**	**.00136**	**-.00217**	**-.00007**	**-.02052**
Stddev	.00223	.00395	.00125	.00151	.00072	.00009	.04242
%RSD	4811.9	27.034	84.817	111.15	33.177	132.09	206.74
#1	-.00198	-.01008	-.00234	.00035	-.00228	-.00011	.02000
#2	-.00031	-.01726	-.00004	.00309	-.00284	.00003	-.06461
#3	.00243	-.01654	-.00202	.00064	-.00141	-.00012	-.01694
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00016**	**-.00005**	**.00100**	**-.00041**	F **-.02087**	**.05390**	**-.00380**
Stddev	.00014	.00044	.00061	.00080	.02667	.11692	.00540
%RSD	89.389	864.20	61.266	196.75	127.82	216.92	141.87
#1	.00015	-.00056	.00069	-.00020	-.05130	-.07225	-.00999
#2	.00002	.00022	.00171	.00027	-.00157	.15862	-.00007
#3	.00030	.00019	.00061	-.00129	-.00972	.07533	-.00135
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass
High Limit					720.00		
Low Limit					-.02000		

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.03363**	**.00221**	**.00046**	**.04075**	**-.00021**	**.00528**	**-.00228**
Stddev	.02463	.00034	.00014	.01938	.00110	.00417	.00173
%RSD	73.257	15.401	29.515	47.560	531.64	78.936	75.715
#1	.00529	.00248	.00060	.03410	.00021	.00120	-.00371
#2	.04992	.00232	.00047	.02557	-.00145	.00953	-.00036
#3	.04568	.00183	.00032	.06258	.00062	.00510	-.00278
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 28, 2016

Sample Name: L1606117901SDL Acquired: 6/27/2016 11:30:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 25 Custom ID2: Custom ID3:
Comment: WG574214-02

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00148**	**.00052**	**-.02939**	**-.00005**	**.00027**	**-.00175**	**.00170**
Stddev	.00281	.00262	.00025	.00021	.00036	.00601	.00037
%RSD	189.85	505.51	.84423	394.19	132.35	344.06	22.015
#1	-.00270	.00158	-.02920	-.00027	.00047	-.00069	.00166
#2	-.00348	-.00246	-.02967	.00015	-.00014	.00367	.00209
#3	.00173	.00244	-.02930	-.00004	.00049	-.00822	.00134
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	**-.00042**	**.00042**	**-.00030**
Stddev	.00063	.00012	.00034
%RSD	149.80	28.720	112.76
#1	-.00114	.00032	-.00062
#2	.00002	.00038	-.00036
#3	-.00014	.00055	.00006
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10431.**	**63091.**	**3111.2**
Stddev	16.	252.	21.2
%RSD	.15804	.39869	.68135
#1	10444.	63211.	3098.3
#2	10438.	63260.	3135.7
#3	10413.	62802.	3099.7

Approved: June 28, 2016

Sample Name: CCV Acquired: 6/27/2016 11:34:45 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.41029**	**10.314**	**.41311**	**.51300**	**1.0292**	**.05084**	**10.180**	**.05150**
Stddev	.00136	.009	.00246	.00180	.0068	.00017	.038	.00051
%RSD	.33182	.09125	.59569	.35017	.65999	.32692	.37198	.99887
#1	.40919	10.310	.41560	.51161	1.0232	.05065	10.141	.05186
#2	.40986	10.307	.41306	.51237	1.0366	.05092	10.217	.05173
#3	.41181	10.325	.41068	.51503	1.0278	.05095	10.182	.05091
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.20710**	**.51032**	**.52099**	**4.0892**	**50.815**	**1.0221**	**10.252**	**.51220**
Stddev	.00166	.00085	.00175	.0040	.318	.0029	.037	.00281
%RSD	.80235	.16676	.33637	.09673	.62496	.28272	.36345	.54826
#1	.20850	.50937	.52285	4.0847	50.552	1.0201	10.295	.50976
#2	.20755	.51102	.52076	4.0921	51.168	1.0255	10.238	.51527
#3	.20526	.51055	.51937	4.0908	50.726	1.0208	10.224	.51156
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.0210**	**51.237**	**.51895**	**10.296**	**.51615**	**1.2226**	**.41042**	**5.0648**
Stddev	.0077	.306	.00211	.030	.00345	.0091	.00397	.0173
%RSD	.75227	.59797	.40704	.29113	.66809	.74330	.96773	.34141
#1	1.0285	50.954	.52109	10.330	.51651	1.2329	.41221	5.0786
#2	1.0215	51.563	.51889	10.286	.51254	1.2195	.41319	5.0704
#3	1.0131	51.194	.51687	10.272	.51941	1.2156	.40587	5.0454
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Approved: June 28, 2016

Sample Name: CCV Acquired: 6/27/2016 11:34:45 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.0301**	**1.0276**	**1.0224**	**.52111**	**1.0183**	**1.0425**	**1.0206**
Stddev	.0050	.0038	.0117	.00210	.0025	.0036	.0032
%RSD	.48419	.37352	1.1487	.40271	.24354	.34135	.30940
#1	1.0346	1.0244	1.0153	.52324	1.0181	1.0453	1.0170
#2	1.0309	1.0319	1.0359	.52105	1.0159	1.0436	1.0231
#3	1.0247	1.0267	1.0159	.51905	1.0209	1.0385	1.0217
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9545.4**	**57355.**	**2920.6**
Stddev	16.0	98.	12.9
%RSD	.16742	.17123	.44036
#1	9535.4	57468.	2935.4
#2	9537.0	57297.	2914.0
#3	9563.8	57299.	2912.3

Approved: June 28, 2016

Sample Name: CCB Acquired: 6/27/2016 11:38:22 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00042**	**.00504**	**-.00201**	**.00368**	**.00256**	**.00009**	**.01404**
Stddev	.00219	.00642	.00250	.00148	.00163	.00009	.04911
%RSD	527.90	127.30	124.48	40.110	63.751	100.94	349.75
#1	.00239	.00398	.00030	.00311	.00407	.00018	.06082
#2	-.00194	.01193	-.00166	.00536	.00276	.00009	.01840
#3	.00080	-.00078	-.00467	.00258	.00083	.00000	-.03710
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00012**	**-.00005**	**.00229**	**.00038**	**.02444**	**.20214**	**-.00177**
Stddev	.00013	.00037	.00132	.00078	.03885	.05544	.00386
%RSD	111.48	696.52	57.782	205.08	158.97	27.426	217.74
#1	-.00023	.00007	.00222	.00101	.05224	.16080	.00022
#2	.00003	-.00047	.00100	.00062	-.01995	.26514	-.00622
#3	-.00016	.00024	.00364	-.00049	.04104	.18047	.00068
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.03023**	**.00028**	**.00420**	**.13919**	**-.00004**	F **.01462**	**.00241**
Stddev	.11455	.00120	.00024	.02048	.00017	.00305	.00207
%RSD	378.91	434.05	5.6517	14.715	417.93	20.859	86.062
#1	-.03575	.00137	.00447	.15454	-.00014	.01761	.00384
#2	-.14193	-.00101	.00411	.11593	.00015	.01475	.00003
#3	.08698	.00047	.00403	.14709	-.00013	.01151	.00335
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						.01000	
Low Limit						-.01000	

Approved: June 28, 2016

Sample Name: CCB Acquired: 6/27/2016 11:38:22 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00518	-.00119	.01136	.00155	.00215	-.00792	.00076
Stddev	.00420	.00357	.00200	.00064	.00047	.00257	.00569
%RSD	81.176	299.26	17.574	41.045	21.906	32.387	751.91
#1	.00995	-.00456	.01333	.00222	.00256	-.00563	-.00271
#2	.00206	.00254	.01141	.00146	.00225	-.01069	-.00235
#3	.00351	-.00155	.00934	.00096	.00164	-.00745	.00733
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.00124	.00038	.00159
Stddev	.00096	.00031	.00019
%RSD	76.994	82.329	11.704
#1	.00222	.00051	.00138
#2	.00031	.00062	.00166
#3	.00119	.00002	.00173
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9358.9	56547.	2794.8
Stddev	20.4	165.	12.6
%RSD	.21850	.29146	.44991
#1	9359.7	56591.	2780.9
#2	9338.0	56684.	2805.3
#3	9378.9	56364.	2798.2

Approved: June 28, 2016

Sample Name: L1606102602 Acquired: 6/27/2016 11:42:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00041**	**.26453**	**-.00232**	**.01137**	**.14455**	**.00168**	**156.80**
Stddev	.00213	.00553	.00315	.00190	.00153	.00005	.86
%RSD	514.79	2.0924	135.85	16.713	1.0603	2.7135	.54868
#1	.00257	.27024	-.00572	.00947	.14575	.00166	157.76
#2	.00036	.26416	-.00174	.01139	.14507	.00165	156.58
#3	-.00169	.25919	.00050	.01327	.14282	.00174	156.08
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00793**	**.29026**	**4.1463**	**.79834**	**18.146**	**10.415**	**.14997**
Stddev	.00013	.00197	.0069	.00546	.096	.012	.00155
%RSD	1.6398	.67726	.16696	.68358	.52983	.11605	1.0303
#1	.00806	.29105	4.1518	.80306	18.217	10.402	.15171
#2	.00780	.29171	4.1487	.79960	18.185	10.426	.14875
#3	.00793	.28803	4.1385	.79237	18.036	10.416	.14945
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**117.35**	**2.7508**	**.04403**	F **463.92**	**11.613**	**.23944**	**.00538**
Stddev	.19	.0034	.00069	5.22	.075	.00408	.00236
%RSD	.16323	.12385	1.5776	1.1254	.64419	1.7056	43.762
#1	117.53	2.7538	.04451	469.94	11.676	.23503	.00518
#2	117.15	2.7516	.04435	460.70	11.634	.24308	.00783
#3	117.37	2.7471	.04323	461.11	11.530	.24022	.00313
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass
High Limit				270.00			
Low Limit				-.50000			

Approved: June 28, 2016

Sample Name: L1606102602 Acquired: 6/27/2016 11:42:20 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	F -.03604	.00167	21.855	-.00030	2.7919	-.02523	.00328
Stddev	.00078	.00175	.048	.00077	.0132	.01080	.00542
%RSD	2.1519	104.78	.21795	256.83	.47315	42.798	165.41
#1	-.03690	.00316	21.910	-.00045	2.8052	-.02849	.00549
#2	-.03581	.00211	21.837	.00053	2.7919	-.03401	-.00290
#3	-.03540	-.00026	21.820	-.00099	2.7788	-.01317	.00724
Check ?	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit	45.000						
Low Limit	-.02000						

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	-.00181	.12543	-.00203
Stddev	.00054	.00077	.00044
%RSD	29.945	.61364	21.462
#1	-.00127	.12603	-.00164
#2	-.00180	.12571	-.00195
#3	-.00236	.12456	-.00250
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9341.9	56331.	2987.1
Stddev	32.3	177.	26.1
%RSD	.34579	.31360	.87222
#1	9359.2	56340.	2957.6
#2	9362.0	56151.	2996.7
#3	9304.7	56503.	3007.0

Approved: June 28, 2016

Sample Name: L1606102604 Acquired: 6/27/2016 11:46:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00081	.51537	-.00135	.04849	.25692	.00002	10.597	.00065
Stddev	.00097	.00417	.00160	.00232	.00117	.00006	.071	.00028
%RSD	119.61	.80856	118.17	4.7769	.45645	334.00	.67436	43.209
#1	.00084	.51600	-.00179	.04829	.25614	.00001	10.599	.00088
#2	.00178	.51092	-.00268	.05091	.25827	-.00004	10.668	.00073
#3	-.00017	.51918	.00042	.04629	.25636	.00009	10.525	.00034
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00143	.00804	.00219	.94506	.57282	.01957	8.3432	.04056
Stddev	.00025	.00118	.00070	.02358	.17090	.00552	.1393	.00132
%RSD	17.728	14.672	31.807	2.4955	29.835	28.200	1.6694	3.2511
#1	.00136	.00921	.00247	.92684	.76429	.02561	8.4862	.03904
#2	.00171	.00806	.00270	.97170	.43570	.01479	8.3352	.04139
#3	.00122	.00685	.00140	.93664	.51849	.01831	8.2080	.04125
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00079	111.22	.01184	.05038	.00139	.00283	-.00142	30.099
Stddev	.00005	.58	.00193	.00219	.00224	.00277	.00860	.182
%RSD	6.3443	.52474	16.326	4.3492	161.71	98.192	607.40	.60408
#1	.00082	110.78	.01373	.04932	.00371	.00258	-.00365	30.191
#2	.00082	111.88	.01193	.05290	.00122	.00018	-.00868	30.217
#3	.00073	111.00	.00987	.04892	-.00077	.00571	.00808	29.890
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606102604 Acquired: 6/27/2016 11:46:14 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00070	.21272	.00487	.00126	.00108	.00766	.00061
Stddev	.00022	.00072	.00311	.00378	.00014	.00024	.00054
%RSD	31.509	.33665	63.798	300.82	12.513	3.1489	89.018
#1	.00046	.21236	.00741	.00517	.00119	.00770	.00081
#2	.00077	.21354	.00140	.00095	.00112	.00788	.00102
#3	.00089	.21225	.00580	-.00236	.00093	.00740	-.00001
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9804.0	59120.	3006.8
Stddev	14.0	114.	4.7
%RSD	.14307	.19206	.15632
#1	9819.7	59017.	3005.4
#2	9792.7	59241.	3002.9
#3	9799.5	59101.	3012.0

Approved: June 28, 2016

Sample Name: L1606102606 Acquired: 6/27/2016 11:50:09 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00012	1.2239	-.00174	.01161	.03727	.00027	5.5157
Stddev	.00080	.0104	.00223	.00250	.00221	.00006	.1070
%RSD	652.80	.85109	128.19	21.558	5.9287	21.417	1.9394
#1	.00092	1.2359	.00079	.01170	.03931	.00028	5.6257
#2	.00014	1.2178	-.00341	.00906	.03758	.00032	5.4120
#3	-.00069	1.2180	-.00260	.01407	.03493	.00021	5.5094
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00041	.01245	.00844	.00357	1.6333	1.2394	.06022
Stddev	.00012	.00034	.00111	.00096	.0776	.0714	.00425
%RSD	29.771	2.7570	13.173	26.915	4.7517	5.7615	7.0639
#1	.00032	.01223	.00839	.00440	1.7197	1.3197	.06065
#2	.00055	.01228	.00957	.00252	1.5695	1.2156	.06423
#3	.00036	.01285	.00735	.00380	1.6107	1.1830	.05576
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	5.8139	.13811	.00206	248.44	.05729	.26089	.00094
Stddev	.1184	.00127	.00026	6.69	.00127	.00873	.00227
%RSD	2.0371	.91838	12.474	2.6922	2.2255	3.3451	240.98
#1	5.8695	.13935	.00185	255.89	.05875	.25082	.00120
#2	5.8943	.13817	.00234	242.94	.05639	.26617	-.00144
#3	5.6779	.13682	.00197	246.50	.05673	.26568	.00307
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Approved: June 28, 2016

Sample Name: L1606102606 Acquired: 6/27/2016 11:50:09 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00049	.00105	F 36.384	.00024	.10851	.01712	-.00030
Stddev	.00500	.00476	.192	.00071	.00323	.00573	.00124
%RSD	1024.8	454.95	.52638	293.29	2.9813	33.469	416.21
#1	.00624	-.00445	36.500	-.00042	.11191	.02347	-.00061
#2	-.00191	.00369	36.488	.00015	.10548	.01232	-.00135
#3	-.00286	.00389	36.163	.00100	.10813	.01557	.00107
Check ?	Chk Pass	Chk Pass	Chk Fail	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit			36.000				
Low Limit			-1.0000				

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.00535	.16936	.00138
Stddev	.00042	.00122	.00039
%RSD	7.7851	.71943	28.035
#1	.00578	.16980	.00152
#2	.00533	.17029	.00168
#3	.00495	.16798	.00094
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9626.3	57886.	2983.8
Stddev	18.3	261.	74.6
%RSD	.19058	.45101	2.5006
#1	9647.5	57585.	2901.3
#2	9616.7	58049.	3046.6
#3	9614.8	58025.	3003.6

Approved: June 28, 2016

Sample Name: L1606102607 Acquired: 6/27/2016 11:54:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00031	1.1323	-.00238	.04660	.10958	.00010	6.5258	.00037
Stddev	.00080	.0175	.00050	.00030	.00243	.00010	.0579	.00010
%RSD	257.95	1.5429	20.924	.64173	2.2156	104.38	.88676	26.723
#1	-.00092	1.1127	-.00240	.04687	.10738	.00021	6.4875	.00037
#2	.00060	1.1463	-.00188	.04628	.10917	.00010	6.4976	.00027
#3	-.00061	1.1378	-.00287	.04666	.11218	-.00000	6.5924	.00047
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00187	.00297	.00159	1.3290	.57069	.00275	3.8665	.01222
Stddev	.00009	.00021	.00046	.0290	.04032	.00285	.0619	.00283
%RSD	4.6234	7.1037	28.704	2.1787	7.0656	103.56	1.6015	23.182
#1	.00187	.00273	.00106	1.3104	.54271	.00511	3.8437	.01540
#2	.00178	.00310	.00180	1.3143	.61691	.00356	3.8192	.00998
#3	.00195	.00309	.00189	1.3624	.55245	-.00042	3.9366	.01126
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00103	20.808	.00171	.02360	-.00106	.00153	.00040	8.8461
Stddev	.00018	.091	.00048	.00800	.00292	.00392	.00140	.1098
%RSD	17.429	.43937	27.803	33.897	275.11	256.50	351.43	1.2414
#1	.00085	20.802	.00191	.01627	-.00308	.00603	-.00121	8.9294
#2	.00103	20.719	.00205	.03214	.00229	-.00117	.00118	8.8872
#3	.00121	20.902	.00117	.02239	-.00240	-.00027	.00122	8.7216
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606102607 Acquired: 6/27/2016 11:54:00 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00017	.12250	.01094	.00018	.00267	.07442	.00061
Stddev	.00032	.00081	.00149	.00374	.00039	.00117	.00047
%RSD	181.48	.65945	13.593	2023.2	14.549	1.5745	76.234
#1	-.00005	.12246	.01133	-.00363	.00276	.07517	.00038
#2	.00054	.12171	.01218	.00033	.00300	.07502	.00115
#3	.00004	.12332	.00929	.00385	.00224	.07307	.00031
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9791.2	59635.	2975.2
Stddev	8.5	18.	25.1
%RSD	.08713	.02982	.84202
#1	9783.0	59652.	2959.7
#2	9790.4	59616.	3004.1
#3	9800.1	59637.	2961.9

Approved: June 28, 2016

Sample Name: L1606084001 Acquired: 6/27/2016 11:57:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00102	.00453	-.00231	.01392	.03067	-.00008	53.120	.00080
Stddev	.00088	.01618	.00071	.00201	.00012	.00007	.441	.00016
%RSD	86.051	357.15	30.592	14.417	.38086	81.244	.83058	20.432
#1	.00104	-.00250	-.00154	.01565	.03054	-.00015	53.330	.00085
#2	.00013	-.00695	-.00246	.01439	.03076	-.00007	53.418	.00062
#3	.00188	.02304	-.00294	.01172	.03072	-.00002	52.613	.00093
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00081	.00170	.00123	.32459	1.0265	-.00206	12.392	.94460
Stddev	.00013	.00072	.00102	.00668	.0613	.00244	.121	.00127
%RSD	16.603	42.489	83.413	2.0577	5.9742	118.69	.97455	.13438
#1	.00096	.00095	.00005	.31956	1.0133	-.00472	12.401	.94427
#2	.00077	.00177	.00186	.33217	.97280	-.00155	12.267	.94353
#3	.00070	.00239	.00178	.32204	1.0933	.00009	12.508	.94600
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00065	5.1723	.00021	.01672	-.00134	-.00196	.00063	4.1375
Stddev	.00018	.0275	.00068	.00105	.00210	.00242	.00481	.0318
%RSD	28.323	.53205	318.26	6.3027	156.43	123.76	761.62	.76821
#1	.00086	5.1913	.00009	.01551	.00023	.00018	.00452	4.1074
#2	.00051	5.1849	-.00040	.01723	-.00053	-.00459	.00212	4.1343
#3	.00057	5.1407	.00095	.01742	-.00372	-.00146	-.00475	4.1707
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606084001 Acquired: 6/27/2016 11:57:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00070**	**.24263**	**-.01886**	**-.00031**	**-.00033**	**.01399**	**-.00001**
Stddev	.00032	.00045	.00470	.00047	.00081	.00026	.00028
%RSD	46.464	.18625	24.915	152.40	243.31	1.8674	2189.7
#1	-.00052	.24229	-.02136	-.00072	-.00123	.01369	-.00029
#2	-.00050	.24315	-.01344	-.00041	.00034	.01408	-.00002
#3	-.00107	.24246	-.02177	.00020	-.00011	.01419	.00027
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9610.5**	**58693.**	**2932.5**
Stddev	88.7	624.	9.6
%RSD	.92300	1.0624	.32835
#1	9677.1	58033.	2921.5
#2	9644.5	59272.	2936.8
#3	9509.8	58775.	2939.2

Approved: June 28, 2016

Sample Name: L1606084002 Acquired: 6/27/2016 12:01:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00128	-.00139	.00224	.08634	.09552	-.00002	120.61	.00014
Stddev	.00250	.00876	.00171	.00338	.00036	.00006	.69	.00026
%RSD	194.67	631.07	76.296	3.9175	.37645	245.45	.57215	192.80
#1	-.00402	-.01109	.00064	.08244	.09586	.00004	119.84	.00033
#2	-.00069	.00593	.00204	.08824	.09555	-.00007	121.15	-.00016
#3	.00087	.00100	.00403	.08835	.09514	-.00004	120.85	.00023
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00063	.00077	.00147	43.081	4.5386	.00310	11.268	.67312
Stddev	.00017	.00005	.00320	.235	.0338	.00167	.145	.00147
%RSD	26.964	6.9558	217.13	.54506	.74492	54.079	1.2833	.21851
#1	-.00075	.00072	.00503	42.809	4.5769	.00120	11.171	.67359
#2	-.00043	.00076	.00055	43.216	4.5263	.00372	11.198	.67430
#3	-.00069	.00082	-.00116	43.216	4.5127	.00437	11.434	.67147
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00116	5.4244	.00012	.11801	.00233	.00183	-.00652	7.6403
Stddev	.00020	.0386	.00089	.00434	.00125	.00095	.00211	.1136
%RSD	17.017	.71102	759.90	3.6803	53.783	51.671	32.377	1.4866
#1	.00110	5.3838	-.00060	.11342	.00098	.00292	-.00818	7.7661
#2	.00100	5.4288	.00111	.11857	.00255	.00144	-.00414	7.6097
#3	.00138	5.4605	-.00016	.12205	.00346	.00115	-.00724	7.5452
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606084002 Acquired: 6/27/2016 12:01:48 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00059**	**.68115**	**.00149**	**.00003**	**.00035**	**.00200**	**-.00075**
Stddev	.00013	.00063	.00809	.00250	.00042	.00020	.00050
%RSD	21.360	.09243	543.95	9778.5	119.63	9.8337	66.242
#1	.00061	.68055	.00036	-.00281	-.00013	.00223	-.00067
#2	.00046	.68181	-.00598	.00096	.00057	.00186	-.00128
#3	.00071	.68110	.01008	.00193	.00063	.00192	-.00029
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9739.0**	**57938.**	**2928.4**
Stddev	102.1	129.	16.1
%RSD	1.0478	.22223	.54818
#1	9622.7	57791.	2946.0
#2	9781.2	57992.	2924.9
#3	9813.3	58031.	2914.5

Approved: June 28, 2016

Sample Name: L1606084003 Acquired: 6/27/2016 12:05:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00121	.08071	.00172	.09081	.09959	-.00012	123.30	.00045
Stddev	.00149	.00384	.00171	.00145	.00117	.00007	1.07	.00020
%RSD	122.93	4.7551	99.258	1.5948	1.1749	56.220	.86783	45.782
#1	.00002	.08463	.00369	.09156	.09973	-.00020	122.33	.00068
#2	.00073	.07696	.00088	.08914	.10068	-.00009	124.45	.00031
#3	.00288	.08054	.00059	.09174	.09835	-.00007	123.11	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00069	.00087	.00289	43.973	4.6960	.00430	11.675	.66419
Stddev	.00015	.00041	.00047	.300	.1683	.00551	.133	.00687
%RSD	21.226	47.747	16.160	.68244	3.5840	127.88	1.1395	1.0342
#1	-.00071	.00059	.00238	43.670	4.7091	-.00204	11.575	.65887
#2	-.00082	.00134	.00298	44.270	4.8574	.00712	11.826	.67195
#3	-.00053	.00067	.00330	43.978	4.5216	.00783	11.625	.66177
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00136	5.5885	.00111	.13500	.00082	.00343	-.00607	7.9482
Stddev	.00015	.0707	.00031	.00464	.00148	.00147	.00397	.0520
%RSD	10.645	1.2649	27.924	3.4356	181.08	42.697	65.305	.65412
#1	.00123	5.5076	.00132	.14035	.00252	.00509	-.00876	7.9966
#2	.00152	5.6384	.00075	.13201	-.00018	.00231	-.00152	7.9546
#3	.00135	5.6196	.00126	.13265	.00012	.00290	-.00794	7.8933
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606084003 Acquired: 6/27/2016 12:05:39 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00054	.69838	.00629	.00178	-.00059	.00222	-.00084
Stddev	.00157	.00675	.00940	.00291	.00056	.00018	.00039
%RSD	291.19	.96596	149.56	163.25	95.098	8.3166	46.054
#1	-.00069	.69240	.00543	.00493	-.00119	.00201	-.00053
#2	.00110	.70570	.01609	.00124	-.00007	.00234	-.00128
#3	-.00203	.69705	-.00266	-.00082	-.00052	.00232	-.00072
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9762.2	58678.	2957.0
Stddev	32.4	138.	19.3
%RSD	.33194	.23584	.65148
#1	9726.5	58612.	2972.9
#2	9770.6	58584.	2935.6
#3	9789.6	58837.	2962.6

Approved: June 28, 2016

Sample Name: L1606084004 Acquired: 6/27/2016 12:09:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00340	.57839	.00309	.04986	.07128	-.00014	78.157	.00024
Stddev	.00231	.00473	.00177	.00231	.00149	.00006	.489	.00016
%RSD	67.902	.81696	57.428	4.6304	2.0875	41.855	.62597	66.282
#1	.00236	.58363	.00106	.04772	.07297	-.00021	78.703	.00036
#2	.00179	.57447	.00437	.05231	.07070	-.00013	77.759	.00006
#3	.00604	.57706	.00382	.04954	.07017	-.00009	78.008	.00030
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00073	.00268	.00209	13.902	2.3115	.00063	10.947	6.5059
Stddev	.00048	.00131	.00044	.065	.1653	.00105	.167	.0351
%RSD	65.715	48.910	20.939	.46649	7.1526	166.23	1.5274	.53929
#1	.00027	.00338	.00167	13.930	2.4961	.00110	10.946	6.5456
#2	.00070	.00349	.00255	13.949	2.1770	.00137	10.779	6.4928
#3	.00124	.00117	.00206	13.828	2.2613	-.00057	11.114	6.4792
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00088	6.4050	.00124	.08236	.00411	-.00373	.00110	7.1748
Stddev	.00021	.0135	.00010	.00949	.00291	.00074	.00066	.0315
%RSD	24.174	.21060	7.7642	11.522	70.779	19.789	59.906	.43949
#1	.00092	6.4182	.00133	.09095	.00714	-.00298	.00137	7.1943
#2	.00065	6.3912	.00123	.08397	.00386	-.00375	.00158	7.1917
#3	.00106	6.4055	.00114	.07217	.00134	-.00446	.00035	7.1384
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606084004 Acquired: 6/27/2016 12:09:30 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00046**	**.46037**	**.01892**	**-.00148**	**.00190**	**.00560**	**.00084**
Stddev	.00044	.00319	.00662	.00110	.00051	.00012	.00070
%RSD	94.473	.69285	34.994	74.711	26.966	2.1834	83.839
#1	-.00082	.46398	.02521	-.00240	.00240	.00568	.00094
#2	.00003	.45792	.01952	-.00177	.00138	.00566	.00149
#3	-.00060	.45920	.01202	-.00026	.00192	.00546	.00009
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9834.1**	**59607.**	**3008.2**
Stddev	10.8	123.	8.6
%RSD	.10933	.20604	.28680
#1	9822.0	59582.	3001.2
#2	9842.6	59499.	3017.8
#3	9837.6	59741.	3005.5

Approved: June 28, 2016

Sample Name: L1606084005 Acquired: 6/27/2016 12:13:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00007	.00668	-.00213	.05778	.13125	-.00012	73.966	.00069
Stddev	.00063	.00810	.00246	.00044	.00094	.00005	.524	.00019
%RSD	961.46	121.37	115.65	.76854	.71462	44.545	.70822	27.315
#1	.00045	.01321	-.00423	.05814	.13220	-.00015	74.552	.00072
#2	.00012	-.00239	-.00273	.05791	.13032	-.00015	73.804	.00085
#3	-.00077	.00921	.00058	.05728	.13121	-.00006	73.542	.00048
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00041	.00184	.00120	7.9376	2.7914	-.00180	9.2146	6.0686
Stddev	.00052	.00037	.00060	.0230	.0535	.00432	.1741	.0460
%RSD	124.87	20.296	50.060	.29024	1.9151	239.98	1.8892	.75859
#1	.00002	.00226	.00162	7.9508	2.7329	-.00167	9.4003	6.1216
#2	.00100	.00155	.00148	7.9110	2.8376	.00245	9.1881	6.0382
#3	.00022	.00171	.00051	7.9510	2.8038	-.00619	9.0552	6.0461
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00401	5.5235	.00009	.03809	.00162	.00315	.00025	5.3185
Stddev	.00034	.0158	.00120	.00432	.00281	.00219	.00692	.0583
%RSD	8.3556	.28595	1268.4	11.330	173.81	69.373	2730.8	1.0955
#1	.00384	5.5412	-.00129	.03382	.00477	.00067	-.00567	5.2569
#2	.00440	5.5110	.00065	.04245	-.00066	.00477	-.00143	5.3727
#3	.00379	5.5181	.00092	.03799	.00075	.00402	.00786	5.3258
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606084005 Acquired: 6/27/2016 12:13:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00032	.43025	-.01337	.00358	.00145	.00273	.00046
Stddev	.00078	.00329	.00402	.00506	.00034	.00007	.00083
%RSD	245.46	.76446	30.110	141.13	23.301	2.6980	182.02
#1	-.00058	.43394	-.01623	-.00199	.00115	.00271	.00009
#2	.00069	.42916	-.00876	.00789	.00138	.00281	-.00013
#3	.00083	.42764	-.01510	.00485	.00182	.00266	.00141
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9841.4	59195.	3025.3
Stddev	94.8	297.	10.0
%RSD	.96296	.50136	.33059
#1	9950.6	59537.	3015.3
#2	9781.1	59011.	3035.3
#3	9792.5	59037.	3025.3

Approved: June 28, 2016

Sample Name: L1606084006 Acquired: 6/27/2016 12:17:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00164	-.00719	-.00434	.00375	-.00054	.00002	.11974	.00043
Stddev	.00261	.00725	.00073	.00189	.00054	.00006	.00960	.00020
%RSD	159.01	100.86	16.872	50.405	100.03	227.54	8.0185	45.897
#1	-.00010	-.00070	-.00454	.00485	-.00005	-.00000	.11624	.00065
#2	.00465	-.00585	-.00353	.00157	-.00045	-.00001	.13060	.00028
#3	.00038	-.01502	-.00495	.00483	-.00112	.00009	.11238	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00015	.00229	.00265	-.00130	.02514	-.00449	.04252	.00721
Stddev	.00026	.00051	.00063	.00266	.08048	.00083	.04044	.00181
%RSD	173.46	22.403	23.726	203.56	320.08	18.531	95.099	25.105
#1	.00029	.00183	.00312	-.00319	.11298	-.00545	.04311	.00535
#2	.00032	.00284	.00289	.00173	-.04506	-.00414	.00179	.00897
#3	-.00015	.00218	.00193	-.00246	.00751	-.00389	.08265	.00733
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00019	.18897	.00020	-.00594	-.00252	-.00088	-.00520	-.00198
Stddev	.00048	.00835	.00100	.00884	.00248	.00457	.00518	.00107
%RSD	253.64	4.4186	510.98	148.75	98.389	518.63	99.733	53.864
#1	-.00019	.19627	-.00072	.00197	-.00532	.00083	-.00935	-.00093
#2	.00002	.17986	.00004	-.01549	-.00162	.00258	.00061	-.00194
#3	.00073	.19077	.00127	-.00432	-.00061	-.00606	-.00686	-.00307
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606084006 Acquired: 6/27/2016 12:17:33 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00085	.00072	-.00045	-.00026	.00094	.00173	-.00031
Stddev	.00035	.00028	.00524	.00154	.00035	.00024	.00032
%RSD	41.281	38.237	1171.4	600.81	37.636	13.582	100.73
#1	.00126	.00102	-.00580	.00038	.00116	.00170	-.00050
#2	.00070	.00048	.00466	-.00202	.00112	.00152	-.00049
#3	.00061	.00067	-.00020	.00086	.00053	.00198	.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9745.8	60648.	2978.8
Stddev	99.8	834.	31.6
%RSD	1.0242	1.3759	1.0602
#1	9633.7	61604.	2998.6
#2	9778.5	60275.	2942.3
#3	9825.2	60065.	2995.3

Approved: June 28, 2016

Sample Name: CCV Acquired: 6/27/2016 12:21:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40982**	**10.326**	**.41213**	**.51242**	**1.0926**	**.05082**	**10.880**	**.05201**
Stddev	.00878	.252	.01063	.01050	.0236	.00138	.223	.00114
%RSD	2.1419	2.4371	2.5798	2.0490	2.1601	2.7195	2.0480	2.1946
#1	.40490	10.191	.41289	.50779	1.1037	.04992	11.005	.05241
#2	.41995	10.616	.40114	.52444	1.0655	.05241	10.623	.05073
#3	.40460	10.170	.42237	.50503	1.1087	.05012	11.012	.05290
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.20803**	**.50919**	**.52134**	**4.3447**	**53.527**	**1.0805**	**10.776**	**.54239**
Stddev	.00484	.01280	.01202	.1267	1.095	.0260	.268	.00962
%RSD	2.3249	2.5137	2.3048	2.9152	2.0452	2.4085	2.4916	1.7730
#1	.20997	.50240	.52340	4.3706	54.097	1.0939	10.816	.54819
#2	.20252	.52396	.50843	4.2071	52.265	1.0505	10.489	.53129
#3	.21159	.50123	.53220	4.4564	54.220	1.0971	11.021	.54769
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.0267**	**54.607**	**.52138**	**10.371**	**.52185**	**1.2269**	**.41575**	**5.0879**
Stddev	.0227	1.123	.01426	.241	.01274	.0284	.01189	.1164
%RSD	2.2055	2.0562	2.7346	2.3230	2.4413	2.3136	2.8599	2.2884
#1	1.0360	55.161	.52713	10.417	.52334	1.2314	.41415	5.1192
#2	1.0009	53.314	.50515	10.110	.50843	1.1965	.40474	4.9590
#3	1.0433	55.345	.53187	10.585	.53378	1.2528	.42836	5.1854
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Approved: June 28, 2016

Sample Name: CCV Acquired: 6/27/2016 12:21:30 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.0336**	**1.0901**	**1.0684**	**.51892**	**1.0206**	**1.0509**	**1.0769**
Stddev	.0240	.0234	.0198	.00908	.0266	.0248	.0112
%RSD	2.3229	2.1457	1.8498	1.7494	2.6067	2.3611	1.0424
#1	1.0429	1.1002	1.0762	.52256	1.0058	1.0571	1.0826
#2	1.0064	1.0633	1.0460	.50859	1.0513	1.0236	1.0640
#3	1.0516	1.1067	1.0831	.52561	1.0046	1.0721	1.0842
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9564.4**	**57358.**	**2776.3**
Stddev	173.7	1171.	41.5
%RSD	1.8165	2.0418	1.4935
#1	9502.6	58080.	2762.2
#2	9760.6	56007.	2823.0
#3	9430.0	57987.	2743.8

Approved: June 28, 2016

Sample Name: CCB Acquired: 6/27/2016 12:25:07 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00018	.00801	-.00103	.00276	.00212	.00010	.01237
Stddev	.00054	.00191	.00555	.00165	.00059	.00005	.02593
%RSD	305.81	23.874	537.57	59.553	27.793	47.886	209.61
#1	-.00026	.00966	.00519	.00338	.00190	.00007	.02126
#2	.00001	.00845	-.00548	.00401	.00279	.00007	.03268
#3	.00077	.00592	-.00281	.00090	.00167	.00015	-.01683
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00009	-.00008	.00176	.00058	-.00953	.07351	-.00534
Stddev	.00020	.00028	.00092	.00111	.01699	.12532	.00361
%RSD	218.62	347.28	52.557	191.12	178.34	170.47	67.695
#1	-.00011	-.00011	.00227	.00172	-.02080	.11336	-.00262
#2	.00029	.00021	.00069	.00051	-.01780	.17407	-.00944
#3	.00010	-.00035	.00231	-.00049	.01002	-.06689	-.00396
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.03981	.00109	.00405	.14258	-.00036	F .01086	-.00169
Stddev	.10294	.00213	.00039	.04484	.00148	.00691	.00374
%RSD	258.57	194.73	9.6251	31.448	414.81	63.624	221.75
#1	.05401	.00259	.00446	.16898	.00116	.01849	-.00130
#2	-.06949	-.00134	.00368	.16795	-.00043	.00505	.00185
#3	.13492	.00203	.00401	.09081	-.00180	.00903	-.00561
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						.01000	
Low Limit						-.01000	

Approved: June 28, 2016

Sample Name: CCB Acquired: 6/27/2016 12:25:07 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00417	-.00073	.01248	.00092	.00175	-.00304	-.00057
Stddev	.00221	.00348	.00144	.00071	.00054	.00973	.00215
%RSD	53.083	478.32	11.507	77.419	30.591	319.59	377.76
#1	.00663	.00318	.01101	.00084	.00217	.00790	.00142
#2	.00236	-.00349	.01254	.00166	.00193	-.00634	-.00285
#3	.00351	-.00187	.01388	.00025	.00115	-.01070	-.00028
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.00062	.00044	.00126
Stddev	.00058	.00021	.00081
%RSD	93.624	46.980	64.473
#1	.00079	.00067	.00217
#2	-.00003	.00035	.00059
#3	.00109	.00029	.00103
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9088.4	55797.	2803.9
Stddev	32.8	363.	16.7
%RSD	.36063	.64995	.59502
#1	9123.1	56185.	2793.2
#2	9058.0	55741.	2823.1
#3	9084.0	55466.	2795.3

Approved: June 28, 2016

Sample Name: L1606099604 Acquired: 6/27/2016 12:29:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00019	.00580	.00190	.08139	.16840	-.00013	103.32	.00028
Stddev	.00128	.00535	.00228	.00254	.00185	.00011	.61	.00017
%RSD	659.73	92.321	120.00	3.1234	1.1007	81.225	.59136	59.820
#1	-.00019	.00495	.00330	.08419	.17029	-.00020	104.00	.00042
#2	.00163	.00092	-.00073	.08076	.16658	-.00019	103.14	.00010
#3	-.00085	.01152	.00315	.07922	.16832	-.00001	102.82	.00032
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00208	.00210	.00220	6.9135	5.7572	.00318	22.797	1.3714
Stddev	.00011	.00106	.00087	.0506	.1417	.00514	.164	.0068
%RSD	5.4227	50.496	39.640	.73143	2.4607	161.87	.71742	.49651
#1	.00195	.00113	.00136	6.9709	5.9167	.00682	22.816	1.3781
#2	.00218	.00193	.00214	6.8940	5.7092	-.00270	22.951	1.3717
#3	.00210	.00323	.00310	6.8755	5.6458	.00541	22.625	1.3644
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00150	62.898	.00398	.00777	.00038	.00178	-.00471	6.2512
Stddev	.00040	.352	.00112	.00125	.00232	.00142	.00311	.1087
%RSD	26.641	.56029	28.045	16.108	606.93	79.640	66.020	1.7380
#1	.00104	63.287	.00289	.00700	.00207	.00015	-.00636	6.1377
#2	.00168	62.806	.00392	.00922	.00135	.00276	-.00666	6.3542
#3	.00178	62.600	.00512	.00710	-.00227	.00244	-.00112	6.2616
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606099604 Acquired: 6/27/2016 12:29:05 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00021**	**.65534**	**-.02030**	**.00255**	**.00033**	**.00193**	**.00063**
Stddev	.00084	.00661	.00061	.00162	.00088	.00009	.00010
%RSD	408.85	1.0090	3.0257	63.401	268.16	4.7892	15.595
#1	.00026	.66296	-.01985	.00143	.00108	.00204	.00054
#2	.00030	.65206	-.02005	.00182	-.00064	.00187	.00073
#3	-.00118	.65102	-.02100	.00441	.00055	.00188	.00063
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9558.4**	**57612.**	**2936.4**
Stddev	116.5	315.	16.4
%RSD	1.2191	.54701	.55800
#1	9690.7	57619.	2920.8
#2	9471.2	57923.	2953.4
#3	9513.3	57293.	2935.0

Approved: June 28, 2016

Sample Name: L1606099605 Acquired: 6/27/2016 12:32:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00062	.00440	.00035	.08600	.15334	-.00004	92.678	.00021
Stddev	.00208	.00968	.00234	.00157	.00545	.00009	2.912	.00015
%RSD	334.34	219.84	676.18	1.8291	3.5523	214.26	3.1425	70.670
#1	-.00176	.01391	-.00219	.08664	.15180	.00006	91.339	.00015
#2	-.00188	.00474	.00081	.08715	.14883	-.00009	90.677	.00010
#3	.00178	-.00544	.00241	.08421	.15939	-.00010	96.020	.00038
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00365	.00086	.00243	6.0484	5.7709	.00063	31.503	.66090
Stddev	.00047	.00042	.00069	.1770	.1598	.00208	1.094	.02168
%RSD	12.764	48.920	28.484	2.9268	2.7695	327.96	3.4720	3.2802
#1	.00403	.00109	.00318	5.9427	5.7426	.00246	30.808	.65326
#2	.00379	.00111	.00182	5.9498	5.6271	.00107	30.938	.64407
#3	.00313	.00037	.00230	6.2528	5.9430	-.00163	32.764	.68536
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00159	61.738	.00237	.01553	-.00133	-.00075	.00297	7.5875
Stddev	.00014	1.868	.00046	.00404	.00513	.00268	.00109	.0444
%RSD	9.1322	3.0264	19.551	26.021	385.98	357.61	36.581	.58550
#1	.00174	60.825	.00241	.01740	.00350	-.00313	.00247	7.6172
#2	.00145	60.502	.00188	.01089	-.00672	.00216	.00422	7.5364
#3	.00158	63.888	.00280	.01829	-.00077	-.00128	.00222	7.6088
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606099605 Acquired: 6/27/2016 12:32:59 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00035	.55768	-.01630	.00067	-.00017	.00329	-.00008
Stddev	.00043	.01691	.00417	.00102	.00046	.00006	.00023
%RSD	123.49	3.0323	25.549	153.35	265.30	1.7576	273.28
#1	-.00015	.55046	-.02083	.00130	.00034	.00334	.00017
#2	.00059	.54557	-.01263	.00121	-.00055	.00330	-.00027
#3	.00062	.57700	-.01545	-.00051	-.00031	.00322	-.00016
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9505.0	58408.	2954.9
Stddev	81.9	1154.	71.1
%RSD	.86161	1.9755	2.4052
#1	9512.0	57442.	2997.1
#2	9583.1	58096.	2994.7
#3	9419.8	59685.	2872.8

Approved: June 28, 2016

Sample Name: L1606099606 Acquired: 6/27/2016 12:36:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00021	-.00165	-.00025	.08564	.14629	-.00000	89.804	.00032
Stddev	.00112	.01113	.00212	.00090	.00530	.00004	2.055	.00020
%RSD	524.99	674.83	853.55	1.0466	3.6261	1039.4	2.2878	63.164
#1	-.00107	-.01404	-.00260	.08522	.14152	-.00005	87.909	.00022
#2	.00100	.00748	.00036	.08503	.14536	.00003	89.515	.00019
#3	.00071	.00161	.00150	.08667	.15200	.00000	91.988	.00055
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00381	.00210	.00165	5.4359	5.6681	.00480	30.776	.64316
Stddev	.00033	.00040	.00014	.1393	.3324	.00308	.659	.01538
%RSD	8.6138	18.872	8.4193	2.5628	5.8637	64.199	2.1405	2.3911
#1	.00416	.00166	.00181	5.3390	5.2912	.00550	30.170	.63168
#2	.00374	.00222	.00155	5.3732	5.9193	.00143	30.682	.63716
#3	.00352	.00242	.00160	5.5956	5.7937	.00748	31.477	.66063
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00113	59.640	.00253	.00574	-.00082	.00120	-.00418	7.2856
Stddev	.00018	1.339	.00036	.00490	.00097	.00213	.00105	.0742
%RSD	16.247	2.2447	14.289	85.380	118.02	177.90	25.030	1.0179
#1	.00127	58.460	.00258	.00918	-.00191	.00329	-.00447	7.3494
#2	.00120	59.366	.00286	.00013	-.00009	.00129	-.00505	7.3032
#3	.00092	61.095	.00214	.00790	-.00046	-.00098	-.00302	7.2042
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606099606 Acquired: 6/27/2016 12:36:53 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00071	.54135	-.01376	.00040	-.00120	.00295	.00019
Stddev	.00026	.01206	.01048	.00370	.00038	.00007	.00031
%RSD	36.596	2.2286	76.185	932.18	31.665	2.5208	166.02
#1	.00092	.53114	-.02355	-.00349	-.00076	.00294	.00053
#2	.00079	.53824	-.01502	.00080	-.00140	.00303	.00011
#3	.00042	.55466	-.00270	.00387	-.00144	.00288	-.00008
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9559.9	57603.	2976.2
Stddev	50.2	882.	41.3
%RSD	.52468	1.5319	1.3891
#1	9523.1	58519.	3014.9
#2	9539.5	56758.	2981.1
#3	9617.0	57532.	2932.6

Approved: June 28, 2016

Sample Name: L1606099607 Acquired: 6/27/2016 12:40:47 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00065	.02230	-.00450	.01108	.02439	-.00009	66.480	.00028
Stddev	.00090	.00160	.00101	.00392	.00092	.00006	.638	.00009
%RSD	138.75	7.1527	22.436	35.403	3.7818	65.215	.95998	33.028
#1	.00137	.02347	-.00529	.01551	.02514	-.00011	66.633	.00038
#2	.00095	.02049	-.00484	.00964	.02466	-.00015	67.029	.00020
#3	-.00037	.02295	-.00336	.00808	.02336	-.00003	65.780	.00026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00008	.00261	.00143	.00596	.78106	-.00208	22.738	.00477
Stddev	.00016	.00097	.00054	.00364	.04317	.00124	.231	.00310
%RSD	209.97	37.223	38.069	61.110	5.5276	59.612	1.0158	65.020
#1	.00010	.00357	.00150	.00834	.79239	-.00102	22.748	.00706
#2	-.00014	.00263	.00194	.00177	.73335	-.00177	22.964	.00601
#3	-.00020	.00163	.00085	.00778	.81744	-.00344	22.502	.00124
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00133	37.923	-.00023	.01749	-.00363	.00068	-.00604	3.0026
Stddev	.00021	.357	.00012	.00610	.00208	.00469	.01001	.0238
%RSD	15.717	.94215	53.708	34.863	57.244	691.11	165.66	.79333
#1	.00125	38.024	-.00022	.01463	-.00430	.00280	.00439	3.0290
#2	.00117	38.219	-.00036	.02449	-.00528	.00393	-.00694	2.9963
#3	.00157	37.526	-.00011	.01335	-.00130	-.00469	-.01558	2.9826
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606099607 Acquired: 6/27/2016 12:40:47 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00021**	**.18785**	**-.01422**	**.00080**	**-.00056**	**.00150**	**.00005**
Stddev	.00064	.00221	.01247	.00073	.00042	.00008	.00019
%RSD	298.26	1.1744	87.668	90.724	75.529	5.2164	388.21
#1	-.00057	.18821	-.01454	-.00004	-.00039	.00152	.00027
#2	-.00060	.18986	-.00160	.00120	-.00025	.00158	-.00007
#3	.00052	.18549	-.02652	.00124	-.00105	.00142	-.00005
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9625.0**	**57772.**	**2941.4**
Stddev	36.2	586.	22.1
%RSD	.37614	1.0139	.75110
#1	9589.4	58436.	2946.1
#2	9661.8	57328.	2917.4
#3	9623.9	57553.	2960.8

Approved: June 28, 2016

Sample Name: L1606099608 Acquired: 6/27/2016 12:44:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00027	-.00400	-.00422	.02442	.03822	-.00001	83.891	.00034
Stddev	.00221	.00770	.00245	.00158	.00186	.00003	2.002	.00031
%RSD	807.22	192.30	58.014	6.4809	4.8795	554.60	2.3863	89.786
#1	.00167	.00126	-.00362	.02390	.03905	.00003	85.735	.00061
#2	.00143	-.01284	-.00691	.02619	.03952	-.00002	84.176	.00001
#3	-.00228	-.00043	-.00212	.02316	.03608	-.00003	81.762	.00040
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00047	.00209	-.00003	-.01556	1.4704	.00084	27.649	.00085
Stddev	.00056	.00083	.00024	.04252	.1367	.00339	.634	.00109
%RSD	120.26	39.828	718.46	273.26	9.2945	402.48	2.2947	127.95
#1	-.00029	.00113	.00015	.01609	1.6204	-.00304	28.230	.00203
#2	-.00110	.00260	.00006	-.06389	1.3530	.00320	27.744	-.00012
#3	-.00002	.00253	-.00031	.00112	1.4378	.00237	26.972	.00065
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00059	14.025	.00030	.00574	-.00047	.00142	-.00271	4.1744
Stddev	.00012	.309	.00096	.00391	.00392	.00326	.00511	.0659
%RSD	19.654	2.2009	321.09	68.078	840.14	229.87	188.88	1.5797
#1	.00046	14.291	-.00049	.00124	.00385	-.00144	-.00858	4.2497
#2	.00065	14.097	.00137	.00774	-.00380	.00496	.00078	4.1468
#3	.00067	13.687	.00002	.00825	-.00145	.00073	-.00032	4.1267
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606099608 Acquired: 6/27/2016 12:44:40 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00084	.27436	-.01978	.00108	-.00021	.00139	.00004
Stddev	.00051	.00737	.00167	.00298	.00020	.00042	.00088
%RSD	60.476	2.6846	8.4580	276.35	96.056	30.038	2176.4
#1	.00048	.28120	-.02080	-.00127	-.00011	.00096	-.00045
#2	.00143	.27533	-.02070	.00007	-.00045	.00179	-.00049
#3	.00062	.26656	-.01785	.00444	-.00007	.00143	.00105
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9658.7	58755.	2964.1
Stddev	91.6	479.	46.2
%RSD	.94862	.81444	1.5591
#1	9554.0	59299.	2924.6
#2	9698.2	58563.	2952.9
#3	9724.0	58402.	3014.9

Approved: June 28, 2016

Sample Name: L1606099609 Acquired: 6/27/2016 12:48:34 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00016	-.00165	-.00164	.02800	.04080	-.00005	79.419	.00020
Stddev	.00144	.01010	.00345	.00084	.00081	.00005	.294	.00013
%RSD	904.18	610.85	210.58	3.0153	1.9890	101.31	.37081	67.386
#1	-.00165	-.00621	.00217	.02847	.04012	-.00005	79.151	.00012
#2	-.00005	-.00866	-.00254	.02851	.04170	-.00010	79.371	.00035
#3	.00122	.00992	-.00454	.02703	.04060	-.00000	79.734	.00012
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00006	.00217	.00166	1.5879	2.0306	-.00104	26.930	.09297
Stddev	.00011	.00067	.00041	.0178	.0539	.00177	.185	.00101
%RSD	198.11	31.130	24.824	1.1212	2.6567	169.28	.68530	1.0817
#1	-.00004	.00198	.00182	1.5969	2.0012	-.00243	26.817	.09181
#2	.00003	.00160	.00198	1.5673	2.0929	.00095	26.831	.09351
#3	.00018	.00291	.00120	1.5993	1.9978	-.00165	27.143	.09359
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00378	15.834	.00047	.00185	-.00068	.00054	-.00357	4.8185
Stddev	.00016	.099	.00142	.00274	.00173	.00291	.00711	.0240
%RSD	4.2873	.62302	302.36	147.91	253.19	536.20	199.32	.49750
#1	.00360	15.733	.00204	.00453	-.00160	.00292	-.01102	4.8453
#2	.00391	15.841	-.00074	.00196	-.00175	.00142	-.00283	4.8113
#3	.00382	15.930	.00011	-.00094	.00131	-.00271	.00315	4.7990
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606099609 Acquired: 6/27/2016 12:48:34 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00041**	**.50689**	**-.01603**	**-.00012**	**.00024**	**.00132**	**.00004**
Stddev	.00051	.00207	.00331	.00296	.00071	.00015	.00003
%RSD	122.28	.40838	20.614	2431.3	295.76	11.101	65.300
#1	-.00086	.50564	-.01656	.00002	.00084	.00147	.00001
#2	-.00052	.50575	-.01250	-.00315	-.00055	.00118	.00004
#3	.00014	.50928	-.01905	.00276	.00043	.00130	.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9626.1**	**58293.**	**2964.2**
Stddev	20.6	405.	2.4
%RSD	.21356	.69449	.08122
#1	9613.0	58609.	2962.4
#2	9649.8	58434.	2966.9
#3	9615.6	57837.	2963.3

Approved: June 28, 2016

Sample Name: L1606111623 Acquired: 6/27/2016 12:52:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00147	.01556	.00438	.34960	.01944	.00464	21.262	.00011
Stddev	.00220	.00714	.00244	.00217	.00139	.00008	.175	.00026
%RSD	150.19	45.914	55.653	.61929	7.1373	1.6984	.82515	246.93
#1	.00261	.02114	.00269	.35205	.01916	.00470	21.452	.00033
#2	-.00107	.01803	.00328	.34794	.02095	.00467	21.229	.00017
#3	.00286	.00751	.00718	.34880	.01822	.00455	21.106	-.00018
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00075	.00359	.00308	.01919	4.9518	.00186	9.2859	.00223
Stddev	.00020	.00096	.00072	.01132	.0871	.00204	.0483	.00253
%RSD	26.204	26.861	23.462	58.996	1.7581	109.65	.51972	113.63
#1	.00052	.00359	.00285	.02309	4.8522	.00180	9.2359	.00360
#2	.00086	.00263	.00390	.00644	4.9899	.00393	9.3323	-.00069
#3	.00087	.00455	.00250	.02805	5.0133	-.00015	9.2894	.00377
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.2853	142.47	.00504	.05255	-.00148	-.00027	.00638	3.3753
Stddev	.0046	.05	.00090	.00619	.00245	.00321	.00461	.0181
%RSD	.36184	.03676	17.858	11.771	165.71	1172.5	72.258	.53501
#1	1.2812	142.46	.00437	.04543	-.00416	.00001	.00525	3.3544
#2	1.2904	142.53	.00606	.05561	-.00095	-.00361	.00244	3.3867
#3	1.2843	142.43	.00468	.05661	.00066	.00278	.01145	3.3846
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606111623 Acquired: 6/27/2016 12:52:27 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00063	.17063	-.00486	.10658	2.7863	.00832	.00293
Stddev	.00042	.00034	.00551	.00239	.0072	.00022	.00040
%RSD	65.420	.19950	113.37	2.2436	.25677	2.6012	13.547
#1	.00097	.17024	-.00043	.10882	2.7881	.00807	.00261
#2	.00017	.17079	-.00312	.10406	2.7784	.00847	.00281
#3	.00077	.17087	-.01102	.10687	2.7923	.00842	.00338
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9640.9	57918.	2931.6
Stddev	63.9	253.	5.2
%RSD	.66309	.43737	.17860
#1	9706.3	57711.	2937.4
#2	9637.9	58200.	2927.2
#3	9578.6	57841.	2930.2

Approved: June 28, 2016

Sample Name: L1606111624 Acquired: 6/27/2016 12:56:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00035	-.00937	.00229	.34841	.01823	.00462	20.849	.00020
Stddev	.00272	.00762	.00155	.00132	.00111	.00006	.357	.00036
%RSD	778.23	81.308	67.611	.37972	6.0899	1.2421	1.7140	175.16
#1	.00231	-.00854	.00224	.34926	.01889	.00469	20.690	.00048
#2	-.00024	-.00220	.00077	.34688	.01695	.00458	20.599	-.00020
#3	-.00312	-.01737	.00386	.34907	.01886	.00460	21.258	.00033
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00094	.00358	.00241	.00141	4.9328	-.00143	8.9933	.00088
Stddev	.00039	.00084	.00071	.01661	.0286	.00191	.2621	.00152
%RSD	41.102	23.367	29.333	1179.3	.58066	133.79	2.9139	171.80
#1	.00128	.00441	.00275	.00238	4.9523	-.00291	8.7732	.00227
#2	.00104	.00274	.00288	.01751	4.9000	.00073	8.9236	.00111
#3	.00052	.00360	.00159	-.01567	4.9463	-.00209	9.2832	-.00074
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.3099	139.68	.00458	.05064	.00100	.00160	.00511	3.3994
Stddev	.0115	1.69	.00119	.00399	.00202	.00134	.01025	.0147
%RSD	.87517	1.2074	26.070	7.8709	201.03	83.535	200.37	.43322
#1	1.3217	138.45	.00428	.05313	.00056	.00152	.01443	3.4160
#2	1.3093	138.98	.00357	.04605	-.00076	.00031	-.00586	3.3942
#3	1.2988	141.60	.00590	.05276	.00320	.00298	.00677	3.3879
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606111624 Acquired: 6/27/2016 12:56:18 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00004**	**.16713**	**-.01305**	**.10902**	**2.8061**	**.00565**	**.00181**
Stddev	.00110	.00200	.00201	.00362	.0198	.00017	.00047
%RSD	2965.9	1.1939	15.396	3.3176	.70560	3.0758	26.191
#1	.00109	.16647	-.01124	.11271	2.8277	.00557	.00141
#2	-.00111	.16554	-.01521	.10886	2.8016	.00554	.00169
#3	-.00010	.16937	-.01269	.10548	2.7889	.00585	.00234
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9621.1**	**58104.**	**2981.3**
Stddev	30.0	389.	29.4
%RSD	.31156	.66925	.98702
#1	9591.7	57688.	2988.3
#2	9651.6	58165.	3006.6
#3	9620.1	58458.	2949.0

Approved: June 28, 2016

Sample Name: L1606102602 Acquired: 6/27/2016 13:00:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00022	.00205	-.00136	.00107	.00338	-.00000	7.7827	.00040
Stddev	.00107	.00744	.00195	.00235	.00045	.00005	.1451	.00005
%RSD	486.55	362.63	143.73	219.19	13.467	1020.0	1.8641	13.006
#1	-.00007	.00972	-.00084	.00337	.00386	-.00001	7.6418	.00039
#2	-.00067	-.00514	.00028	.00118	.00331	.00005	7.9316	.00045
#3	.00141	.00158	-.00351	-.00133	.00296	-.00005	7.7745	.00035
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.01549	.21171	.04406	.85896	.46778	.00399	5.8381	.13373
Stddev	.00042	.00465	.00090	.03523	.09590	.00226	.2053	.00527
%RSD	2.6989	2.1945	2.0404	4.1009	20.502	56.710	3.5163	3.9399
#1	.01595	.21640	.04347	.85962	.49628	.00152	5.6471	.12924
#2	.01537	.20711	.04509	.89386	.36085	.00448	6.0552	.13953
#3	.01514	.21162	.04361	.82342	.54620	.00596	5.8120	.13242
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00408	23.497	.65304	.01847	-.00152	-.00473	.00137	1.1130
Stddev	.00023	.589	.00091	.00293	.00244	.00451	.00512	.0018
%RSD	5.7186	2.5087	.13863	15.888	160.52	95.318	373.36	.15722
#1	.00432	22.966	.65370	.01509	-.00185	-.00890	.00472	1.1127
#2	.00405	24.132	.65201	.01989	.00107	-.00537	.00391	1.1149
#3	.00386	23.394	.65342	.02042	-.00378	.00006	-.00452	1.1114
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606102602 Acquired: 6/27/2016 13:00:08 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00017	.13845	-.00360	.00164	.00318	.00781	-.00040
Stddev	.00028	.00342	.01364	.00088	.00009	.00014	.00006
%RSD	162.07	2.4708	378.91	53.643	2.7914	1.8499	16.297
#1	-.00041	.13497	.01208	.00068	.00319	.00794	-.00047
#2	.00013	.14180	-.01270	.00183	.00309	.00785	-.00038
#3	-.00024	.13857	-.01018	.00241	.00326	.00765	-.00034
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10522.	63691.	3160.7
Stddev	147.	1450.	56.7
%RSD	1.3981	2.2767	1.7935
#1	10388.	62026.	3208.2
#2	10501.	64677.	3097.9
#3	10679.	64370.	3176.0

Approved: June 28, 2016

Sample Name: L1606102604 Acquired: 6/27/2016 13:03:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00120	.02320	-.00068	.00317	.01081	-.00005	.53402	.00002
Stddev	.00006	.00903	.00370	.00080	.00173	.00015	.04625	.00007
%RSD	4.8840	38.933	545.28	25.090	16.044	287.43	8.6606	360.15
#1	.00115	.03309	.00318	.00228	.01279	-.00004	.54280	.00007
#2	.00120	.01539	-.00419	.00343	.01009	.00009	.57525	.00005
#3	.00127	.02113	-.00102	.00380	.00955	-.00021	.48401	-.00006
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00010	.00131	-.00136	.02682	.04953	-.00210	.41888	.00169
Stddev	.00009	.00024	.00075	.02522	.04733	.00367	.07583	.00236
%RSD	89.070	18.283	55.333	94.024	95.563	174.33	18.102	140.13
#1	-.00016	.00148	-.00209	.01757	.00624	-.00538	.50346	.00180
#2	.00000	.00141	-.00059	.00754	.04227	.00186	.35699	.00399
#3	-.00016	.00104	-.00138	.05537	.10006	-.00279	.39619	-.00073
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00091	5.8344	.00024	.00937	-.00078	-.00008	-.00087	1.5329
Stddev	.00017	.1289	.00146	.00288	.00160	.00208	.00351	.0177
%RSD	18.772	2.2091	603.01	30.760	204.78	2500.1	404.55	1.1558
#1	.00104	5.9515	.00088	.00698	-.00120	.00230	-.00352	1.5130
#2	.00098	5.6963	-.00143	.01257	.00098	-.00158	.00311	1.5385
#3	.00072	5.8554	.00127	.00855	-.00212	-.00096	-.00220	1.5471
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606102604 Acquired: 6/27/2016 13:03:58 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00010**	**.01129**	**-.00425**	**.00453**	**.00073**	**.00062**	**-.00033**
Stddev	.00016	.00049	.00124	.00218	.00029	.00001	.00040
%RSD	168.17	4.3032	29.246	48.226	39.627	1.7802	124.10
#1	-.00024	.01184	-.00555	.00662	.00044	.00062	-.00011
#2	.00008	.01092	-.00308	.00226	.00073	.00061	-.00008
#3	-.00014	.01111	-.00410	.00470	.00102	.00063	-.00079
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10264.**	**61770.**	**3063.1**
Stddev	82.	1823.	36.6
%RSD	.80048	2.9515	1.1947
#1	10355.	59854.	3033.0
#2	10241.	61974.	3103.8
#3	10196.	63483.	3052.5

Approved: June 28, 2016

Sample Name: CCV Acquired: 6/27/2016 13:07:53 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39855	9.9772	.42046	.50344	.95772	.04895	9.5448	.05172
Stddev	.00683	.1980	.00527	.01093	.00739	.00095	.0947	.00030
%RSD	1.7128	1.9841	1.2532	2.1714	.77178	1.9362	.99206	.57149
#1	.39069	9.7501	.42628	.49084	.95811	.04786	9.4971	.05193
#2	.40204	10.069	.41910	.50905	.95015	.04942	9.4833	.05186
#3	.40294	10.113	.41601	.51043	.96492	.04957	9.6538	.05138
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20728	.49174	.52427	3.7658	47.113	.95130	9.4667	.47419
Stddev	.00145	.01084	.00292	.0457	.314	.00824	.2111	.00912
%RSD	.70027	2.2054	.55705	1.2121	.66610	.86617	2.2300	1.9228
#1	.20870	.48004	.52679	3.7302	47.107	.95111	9.2849	.48066
#2	.20735	.49371	.52495	3.7499	46.802	.94315	9.4170	.46377
#3	.20580	.50146	.52107	3.8173	47.430	.95963	9.6982	.47816
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.0270	48.116	.52240	10.440	.51931	1.2424	.42085	5.1300
Stddev	.0100	.448	.00357	.065	.00444	.0023	.00479	.0296
%RSD	.97214	.93093	.68267	.62704	.85492	.18669	1.1372	.57801
#1	1.0357	48.131	.52517	10.496	.52062	1.2397	.42531	5.1539
#2	1.0293	47.660	.52366	10.455	.52295	1.2439	.42146	5.1394
#3	1.0161	48.555	.51838	10.368	.51436	1.2434	.41579	5.0968
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Approved: June 28, 2016

Sample Name: CCV Acquired: 6/27/2016 13:07:53 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.0297**	**.95746**	**.93834**	**.51847**	**.98130**	**1.0528**	**.95999**
Stddev	.0063	.00952	.01582	.00063	.02074	.0067	.01007
%RSD	.61354	.99395	1.6856	.12159	2.1136	.63322	1.0495
#1	1.0359	.95765	.94753	.51819	.95739	1.0591	.95621
#2	1.0300	.94785	.92008	.51919	.99197	1.0533	.95235
#3	1.0233	.96688	.94742	.51802	.99452	1.0458	.97141
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9551.0**	**58673.**	**3085.7**
Stddev	21.5	957.	26.3
%RSD	.22509	1.6314	.85238
#1	9526.7	59766.	3082.4
#2	9558.6	58271.	3113.4
#3	9567.7	57983.	3061.2

Approved: June 28, 2016

Sample Name: CCB Acquired: 6/27/2016 13:11:30 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00014**	**-.00445**	**-.00119**	**.00208**	**-.00135**	**-.00001**	**-.01457**	**.00003**
Stddev	.00236	.00341	.00067	.00250	.00099	.00005	.05566	.00031
%RSD	1640.9	76.698	56.614	120.45	73.479	318.32	382.09	1088.0
#1	-.00287	-.00801	-.00141	.00100	-.00037	-.00003	.02640	.00026
#2	.00127	-.00415	-.00043	.00029	-.00133	.00004	-.07794	.00015
#3	.00117	-.00120	-.00172	.00493	-.00235	-.00005	.00784	-.00032
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00027**	**.00199**	**-.00040**	**-.00353**	**.17000**	**-.00171**	**.00538**	**.00159**
Stddev	.00010	.00055	.00131	.03844	.02716	.00277	.06978	.00187
%RSD	35.764	27.862	324.55	1087.7	15.978	162.37	1297.3	117.63
#1	.00032	.00216	.00054	.04081	.15267	-.00456	-.00664	.00232
#2	.00016	.00244	-.00190	-.02729	.20130	.00098	-.05761	.00298
#3	.00034	.00137	.00015	-.02412	.15602	-.00155	.08039	-.00053
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00379**	**.14492**	**.00033**	**.00576**	**-.00060**	**.00422**	**.00182**	**.00473**
Stddev	.00034	.01572	.00035	.01026	.00051	.00038	.00185	.00216
%RSD	8.8594	10.849	107.73	177.93	84.779	8.9533	101.57	45.720
#1	.00391	.14066	.00063	-.00473	-.00071	.00432	.00390	.00674
#2	.00406	.16233	-.00006	.00626	-.00004	.00454	.00119	.00499
#3	.00342	.13177	.00041	.01577	-.00103	.00380	.00036	.00244
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: CCB Acquired: 6/27/2016 13:11:30 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00075	.00183	-.00426	.00298	.00111	.00029	.00146
Stddev	.00071	.00019	.00560	.00028	.00120	.00009	.00044
%RSD	94.654	10.537	131.40	9.4307	107.69	30.310	30.073
#1	-.00003	.00189	-.00424	.00272	.00097	.00036	.00109
#2	.00137	.00161	.00133	.00328	.00238	.00032	.00194
#3	.00091	.00198	-.00988	.00294	-.00001	.00019	.00134
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9906.1	59989.	3002.2
Stddev	128.4	1165.	36.0
%RSD	1.2959	1.9425	1.1986
#1	9929.8	60699.	2993.2
#2	9767.5	58645.	2971.6
#3	10021.	60625.	3041.8

Approved: June 28, 2016

Sample Name: L1606102606 Acquired: 6/27/2016 13:15:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00116	.04310	-.00091	.00135	-.00138	-.00011	.24063	-.00005
Stddev	.00111	.00627	.00105	.00334	.00056	.00005	.02008	.00033
%RSD	96.010	14.536	114.92	247.26	40.605	47.397	8.3444	670.72
#1	.00206	.03718	.00029	.00008	-.00171	-.00017	.26380	-.00039
#2	-.00008	.04246	-.00143	.00513	-.00073	-.00009	.22972	.00028
#3	.00149	.04966	-.00159	-.00117	-.00170	-.00007	.22837	-.00004
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00041	.00158	.00015	.07506	.12658	-.00200	.31903	.00533
Stddev	.00016	.00116	.00212	.01437	.08050	.00181	.11842	.00296
%RSD	39.271	73.288	1442.2	19.147	63.596	90.890	37.118	55.433
#1	.00059	.00233	.00165	.06693	.20692	-.00396	.40405	.00527
#2	.00032	.00216	.00107	.09165	.12692	-.00037	.18378	.00241
#3	.00031	.00025	-.00228	.06659	.04591	-.00166	.36926	.00832
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00102	12.138	.00292	.01735	-.00104	-.00224	-.00069	1.7912
Stddev	.00024	.512	.00070	.00389	.00235	.00340	.00256	.0299
%RSD	23.803	4.2149	24.131	22.406	225.03	151.77	368.24	1.6684
#1	.00119	12.592	.00310	.01880	-.00327	.00006	.00221	1.7616
#2	.00113	11.584	.00351	.02030	-.00127	-.00063	-.00261	1.7906
#3	.00074	12.237	.00214	.01294	.00141	-.00615	-.00168	1.8214
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606102606 Acquired: 6/27/2016 13:15:28 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00038**	**.00523**	**-.00029**	**.00262**	**.00035**	**.00896**	**.00010**
Stddev	.00061	.00027	.00252	.00323	.00155	.00029	.00031
%RSD	158.35	5.1856	883.43	123.20	444.62	3.2537	301.10
#1	-.00031	.00523	.00262	.00309	-.00004	.00869	.00034
#2	-.00102	.00496	-.00161	-.00082	-.00097	.00891	.00022
#3	.00019	.00550	-.00187	.00560	.00206	.00927	-.00025
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**10417.**	**62883.**	**3149.1**
Stddev	112.	932.	103.0
%RSD	1.0779	1.4828	3.2700
#1	10511.	63478.	3064.6
#2	10447.	63364.	3263.8
#3	10293.	61809.	3118.8

Approved: June 28, 2016

Sample Name: L1606129003 Acquired: 6/27/2016 13:19:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.00007	-.00490	-.00263	.00116	.01488	-.00005	.94835	.00015
Stddev	.00069	.00784	.00207	.00095	.00153	.00008	.04705	.00024
%RSD	1049.5	159.97	78.911	82.431	10.267	159.57	4.9610	155.58
#1	.00007	.00408	-.00489	.00009	.01319	-.00005	.89413	.00038
#2	-.00081	-.01036	-.00082	.00191	.01529	-.00014	.97836	-.00009
#3	.00054	-.00843	-.00217	.00147	.01616	.00003	.97256	.00016
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00020	.00029	-.00021	1.5518	.25213	-.00855	.61811	.02470
Stddev	.00039	.00089	.00104	.0399	.11756	.00244	.10475	.00149
%RSD	197.27	311.51	508.66	2.5733	46.628	28.589	16.946	6.0350
#1	.00030	-.00038	.00014	1.5114	.14886	-.01109	.56411	.02301
#2	.00052	-.00006	-.00138	1.5912	.22746	-.00621	.73884	.02584
#3	-.00023	.00130	.00062	1.5527	.38007	-.00834	.55139	.02525
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00028	5.6687	.00119	.01042	-.00074	.00208	.00052	.41967
Stddev	.00026	.1634	.00066	.00843	.00287	.00220	.00352	.01113
%RSD	90.448	2.8821	55.491	80.953	390.80	105.66	671.07	2.6529
#1	.00015	5.4801	.00181	.00715	.00094	.00366	-.00340	.42637
#2	.00058	5.7661	.00050	.00411	-.00406	-.00043	.00157	.42582
#3	.00012	5.7600	.00125	.02000	.00091	.00301	.00340	.40682
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: L1606129003 Acquired: 6/27/2016 13:19:23 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: 20 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00032	.02183	.00273	.00216	.00029	.00131	-.00013
Stddev	.00047	.00102	.00597	.00367	.00048	.00011	.00029
%RSD	147.15	4.6769	218.71	169.95	165.96	8.6066	231.95
#1	.00006	.02065	.00326	.00621	.00084	.00127	.00001
#2	.00004	.02241	-.00349	-.00095	.00002	.00123	-.00046
#3	.00086	.02242	.00842	.00122	.00000	.00144	.00007
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	10192.	62198.	3074.6
Stddev	151.	896.	60.3
%RSD	1.4841	1.4413	1.9622
#1	10087.	63037.	3142.4
#2	10123.	62304.	3026.8
#3	10365.	61254.	3054.7

Approved: June 28, 2016

Sample Name: CCV Acquired: 6/27/2016 13:23:20 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.39466	9.8952	.40043	.50143	.97919	.04849	9.7927	.05038
Stddev	.01306	.3836	.01000	.02169	.00739	.00183	.0660	.00133
%RSD	3.3098	3.8763	2.4963	4.3256	.75498	3.7815	.67452	2.6465
#1	.38704	9.6439	.40978	.48959	.97741	.04728	9.8281	.05184
#2	.38719	9.7049	.40160	.48825	.98731	.04758	9.8334	.05005
#3	.40974	10.337	.38990	.52647	.97285	.05060	9.7165	.04923
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.20069	.48650	.50460	3.8975	47.776	.97248	9.4955	.48926
Stddev	.00441	.01900	.01134	.0630	.388	.00867	.0671	.00174
%RSD	2.1979	3.9054	2.2474	1.6171	.81251	.89105	.70691	.35652
#1	.20508	.47470	.51551	3.8721	47.744	.97251	9.4735	.48919
#2	.20071	.47638	.50543	3.9692	48.180	.98114	9.5709	.49104
#3	.19626	.50842	.49287	3.8510	47.405	.96381	9.4422	.48755
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.99535	49.328	.50566	10.124	.50743	1.2005	.40926	4.9655
Stddev	.02398	.381	.01095	.218	.01442	.0269	.00502	.1016
%RSD	2.4097	.77333	2.1647	2.1554	2.8422	2.2416	1.2268	2.0468
#1	1.0203	49.197	.51675	10.351	.52175	1.2284	.41424	5.0701
#2	.99336	49.758	.50537	10.104	.50763	1.1982	.40933	4.9591
#3	.97242	49.030	.49486	9.9158	.49291	1.1748	.40420	4.8672
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value								
Range								

Approved: June 28, 2016

Sample Name: CCV Acquired: 6/27/2016 13:23:20 Type: QC
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.99630	.97913	.95380	.50300	.97315	1.0214	.97706
Stddev	.02300	.00787	.00906	.01053	.03812	.0240	.00719
%RSD	2.3082	.80329	.94947	2.0942	3.9168	2.3493	.73581
#1	1.0198	.97521	.96013	.51368	.94863	1.0450	.97116
#2	.99534	.98818	.95783	.50269	.95376	1.0222	.97495
#3	.97380	.97399	.94342	.49262	1.0171	.99699	.98507
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
Value							
Range							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9660.1	58486.	3001.0
Stddev	163.9	1564.	9.9
%RSD	1.6964	2.6746	.32945
#1	9505.7	59464.	2995.2
#2	9642.7	59312.	2995.4
#3	9832.0	56682.	3012.5

Approved: June 28, 2016

Sample Name: CCB Acquired: 6/27/2016 13:26:46 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.00010**	**.00303**	**-.00373**	**.00500**	**-.00055**	**-.00003**	**.00393**
Stddev	.00170	.00499	.00268	.00031	.00065	.00008	.03122
%RSD	1689.8	164.71	71.961	6.1638	118.06	299.30	794.56
#1	.00106	.00037	-.00375	.00491	-.00130	.00001	-.03094
#2	.00068	-.00007	-.00103	.00474	-.00008	.00003	.01342
#3	-.00205	.00879	-.00640	.00534	-.00028	-.00012	.02931
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Cd2288	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.00015**	**.00015**	**.00156**	**-.00015**	**-.01189**	**.13544**	**-.00177**
Stddev	.00011	.00034	.00096	.00093	.01157	.09102	.00528
%RSD	70.413	226.11	61.816	600.89	97.339	67.203	297.58
#1	.00003	.00039	.00064	-.00100	-.00558	.20922	-.00741
#2	.00018	-.00024	.00149	-.00030	-.02525	.03373	.00306
#3	.00025	.00029	.00256	.00084	-.00484	.16338	-.00097
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	Mg2790	Mn2576	Mo2020	Na5895	Ni2316	P_2149	Pb2203
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.01205**	**.00050**	**.00395**	**.16416**	**.00017**	F **.01114**	**-.00270**
Stddev	.01807	.00205	.00042	.03616	.00046	.00174	.00166
%RSD	150.01	409.81	10.590	22.026	265.71	15.573	61.523
#1	.02884	.00064	.00398	.19851	-.00024	.01014	-.00178
#2	-.00708	.00248	.00436	.12643	.00066	.01014	-.00462
#3	.01438	-.00162	.00352	.16753	.00010	.01315	-.00170
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Fail	Chk Pass
High Limit						.01000	
Low Limit						-.01000	

Approved: June 28, 2016

Sample Name: CCB Acquired: 6/27/2016 13:26:46 Type: Blank
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sb2068	Se1960	Si2124	Sn1899	Sr4077	Ti3372	Tl1908
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00393	.00112	.00872	.00073	.00181	.01167	.00102
Stddev	.00151	.00746	.00139	.00019	.00058	.00506	.00117
%RSD	38.515	664.39	15.940	25.365	32.293	43.335	115.09
#1	.00316	-.00100	.00967	.00052	.00245	.00931	.00206
#2	.00567	.00941	.00935	.00088	.00131	.00823	.00126
#3	.00295	-.00504	.00712	.00078	.00166	.01748	-.00026
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Elem	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm
Avg	.00126	.00021	.00147
Stddev	.00076	.00021	.00080
%RSD	60.083	103.19	54.681
#1	.00163	.00042	.00075
#2	.00039	.00020	.00233
#3	.00175	-.00000	.00132
Check ?	Chk Pass	Chk Pass	Chk Pass
High Limit			
Low Limit			

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	9947.5	59758.	3010.4
Stddev	96.7	1863.	52.9
%RSD	.97166	3.1170	1.7571
#1	10052.	59475.	3024.7
#2	9862.1	61746.	2951.8
#3	9928.1	58053.	3054.7

Approved: June 28, 2016

Sample Name: LLCCV Acquired: 6/27/2016 13:30:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Ag3280	Al3082	As1890	B_2496	Ba4554	Be3131	Ca4226	Cd2288
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00878	.16766	.00389	.08557	.00677	.00157	.40977	.00073
Stddev	.00207	.00346	.00021	.00327	.00081	.00006	.02595	.00024
%RSD	23.531	2.0622	5.4293	3.8191	11.914	3.9744	6.3339	33.165
#1	.01068	.16888	.00391	.08931	.00721	.00156	.38214	.00070
#2	.00908	.17034	.00409	.08409	.00584	.00164	.41351	.00098
#3	.00658	.16376	.00367	.08329	.00727	.00151	.43365	.00050
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Co2286	Cr2677	Cu2247	Fe2611	K_7664	Li6707	Mg2790	Mn2576
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00384	.00587	.00477	.06082	.91229	.08113	.39200	.00719
Stddev	.00038	.00131	.00041	.02730	.02861	.00171	.10933	.00172
%RSD	9.8067	22.365	8.5727	44.890	3.1359	2.1106	27.890	23.950
#1	.00418	.00493	.00504	.05726	.93538	.08181	.26640	.00586
#2	.00392	.00737	.00430	.03548	.92120	.07919	.44374	.00914
#3	.00343	.00531	.00496	.08973	.88029	.08240	.46585	.00658
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Elem	Mo2020	Na5895	Ni2316	P_2149	Pb2203	Sb2068	Se1960	Si2124
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.00867	.49948	.01573	.80363	.00797	.08206	.01475	.83298
Stddev	.00028	.02628	.00123	.00471	.00119	.00489	.00371	.00454
%RSD	3.1854	5.2609	7.8331	.58570	14.963	5.9538	25.165	.54557
#1	.00835	.47045	.01523	.80835	.00800	.08756	.01866	.83750
#2	.00881	.52163	.01713	.80362	.00915	.08036	.01127	.83302
#3	.00884	.50635	.01482	.79893	.00677	.07824	.01432	.82841
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit								
Low Limit								

Approved: June 28, 2016

Sample Name: LLCCV Acquired: 6/27/2016 13:30:45 Type: Unk
Method: ICP-THERMO3_6010_200.7WATER_3YLINES(v977) Mode: CONC Corr. Factor: 1.00000(
User: JYH Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Sr4077	Ti3372	Tl1908	V_2924	Zn2062	Zr3391
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.40284**	**.04313**	**.02333**	**.15595**	**.00848**	**.01704**	**.04589**
Stddev	.00545	.00068	.00788	.00367	.00039	.00004	.00162
%RSD	1.3539	1.5742	33.792	2.3562	4.5484	.25639	3.5272
#1	.40779	.04330	.03084	.15591	.00827	.01709	.04451
#2	.40373	.04238	.02402	.15965	.00892	.01703	.04548
#3	.39699	.04370	.01512	.15230	.00824	.01701	.04767
Check ?	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass	Chk Pass
High Limit							
Low Limit							

Int. Std.	Y_2243	Y_3600	Y_3774
Units	Cts/S	Cts/S	Cts/S
Avg	**9797.4**	**58651.**	**2881.8**
Stddev	19.7	674.	40.9
%RSD	.20082	1.1484	1.4200
#1	9782.0	58269.	2909.3
#2	9819.6	58255.	2901.4
#3	9790.6	59429.	2834.8

Approved: June 28, 2016

2.3.2 Metals ICP-MS Data

2.3.2.1 Summary Data

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574572	**Analyst:**	BKT	**Run Date:**	06/29/2016 14:51
Collect Date:	06/16/2016 07:55	**Dilution:**	1	**File ID:**	NI.062916.145120
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00134	J	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00894		0.00200	0.00100	0.000500
Barium, Total	7440-39-3	0.186		0.00600	0.00300	0.00150
Cadmium, Total	7440-43-9	0.00645		0.00120	0.000600	0.000300
Lead, Total	7439-92-1	0.00100	U	0.00200	0.00100	0.000500
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Zinc, Total	7440-66-6	0.158		0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:05
Collect Date:	06/16/2016 07:55	**Dilution:**	20	**File ID:**	NI.070516.160533
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Cobalt, Total	7440-48-4	0.355		0.0400	0.0200	0.0100
Copper, Total	7440-50-8	0.828		0.0800	0.0400	0.0200
Manganese, Total	7439-96-5	3.12		0.0800	0.0400	0.0200
Vanadium, Total	7440-62-2	0.0200	U	0.0400	0.0200	0.0100

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/06/2016 09:17
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/06/2016 15:27
Collect Date:	06/16/2016 07:55	**Dilution:**	200	**File ID:**	NI.070616.152752
Sample Tag:	DL03	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Chromium, Total	7440-47-3	4.62		0.800	0.400	0.200

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-02	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	18WW16F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:08
Collect Date:	06/16/2016 07:55	**Dilution:**	200	**File ID:**	NI.070516.160845
Sample Tag:	DL02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Nickel, Total	7440-02-0	14.3		1.60	0.800	0.400

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	129-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574572	**Analyst:**	BKT	**Run Date:**	06/29/2016 14:54
Collect Date:	06/16/2016 09:05	**Dilution:**	1	**File ID:**	NI.062916.145431
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00110	J	0.00200	0.00100	0.000500
Cadmium, Total	7440-43-9	0.000383	J	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00415		0.00400	0.00200	0.00100
Copper, Total	7440-50-8	0.00173	J	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.000875	J	0.00200	0.00100	0.000500
Manganese, Total	7439-96-5	0.0478		0.00400	0.00200	0.00100
Nickel, Total	7440-02-0	0.00956		0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.00246		0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0250	U	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	129-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:11
Collect Date:	06/16/2016 09:05	**Dilution:**	1	**File ID:**	NI.070516.161156
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Cobalt, Total	7440-48-4	0.00182	J	0.00200	0.00100	0.000500

J	Estimated value ; the analyte concentration was less than the LOQ.
J	Estimated value ; the analyte concentration was greater than the highest standard
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-04	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	129-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:15
Collect Date:	06/16/2016 09:05	**Dilution:**	20	**File ID:**	NI.070516.161508
Sample Tag:	DL01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Barium, Total	7440-39-3	0.293		0.120	0.0600	0.0300

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-06	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	120F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574572	**Analyst:**	BKT	**Run Date:**	06/29/2016 14:57
Collect Date:	06/16/2016 09:25	**Dilution:**	1	**File ID:**	NI.062916.145742
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.00356		0.00200	0.00100	0.000500
Barium, Total	7440-39-3	0.0452		0.00600	0.00300	0.00150
Cadmium, Total	7440-43-9	0.000600	U	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.0120		0.00400	0.00200	0.00100
Copper, Total	7440-50-8	0.00476		0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00162	J	0.00200	0.00100	0.000500
Manganese, Total	7439-96-5	0.186		0.00400	0.00200	0.00100
Nickel, Total	7440-02-0	0.0755		0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.0105		0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0659		0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-06	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	120F-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:30
Collect Date:	06/16/2016 09:25	**Dilution:**	1	**File ID:**	NI.070516.163034
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Cobalt, Total	7440-48-4	0.0151		0.00200	0.00100	0.000500

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	123-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	07/05/2016 09:36
Workgroup #:	WG574572	**Analyst:**	JYH	**Run Date:**	07/05/2016 16:34
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	NI.070516.163440
Sample Tag:	02	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Cobalt, Total	7440-48-4	0.00250		0.00200	0.00100	0.000500

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

Certificate of Analysis

Sample #:	L16061026-07	**PrePrep Method:**	N/A	**Instrument:**	ICP-MS2
Client ID:	123-061616	**Prep Method:**	3015	**Prep Date:**	06/21/2016 11:18
Matrix:	Water	**Analytical Method:**	6020A	**Cal Date:**	06/29/2016 11:35
Workgroup #:	WG574572	**Analyst:**	BKT	**Run Date:**	06/29/2016 15:00
Collect Date:	06/16/2016 09:55	**Dilution:**	1	**File ID:**	NI.062916.150054
Sample Tag:	01	**Units:**	mg/L		

Analyte	CAS #	Result	Qual	LOQ	LOD	DL
Antimony, Total	7440-36-0	0.00100	U	0.00200	0.00100	0.000500
Arsenic, Total	7440-38-2	0.000767	J	0.00200	0.00100	0.000500
Barium, Total	7440-39-3	0.136		0.00600	0.00300	0.00150
Cadmium, Total	7440-43-9	0.000600	U	0.00120	0.000600	0.000300
Chromium, Total	7440-47-3	0.00177	J	0.00400	0.00200	0.00100
Copper, Total	7440-50-8	0.00142	J	0.00400	0.00200	0.00100
Lead, Total	7439-92-1	0.00104	J	0.00200	0.00100	0.000500
Manganese, Total	7439-96-5	0.0161		0.00400	0.00200	0.00100
Nickel, Total	7440-02-0	0.00319	J	0.00800	0.00400	0.00200
Silver, Total	7440-22-4	0.00100	U	0.00200	0.00100	0.000500
Thallium, Total	7440-28-0	0.000200	U	0.000400	0.000200	0.000100
Vanadium, Total	7440-62-2	0.00241		0.00200	0.00100	0.000500
Zinc, Total	7440-66-6	0.0250	U	0.0500	0.0250	0.0125

J	Estimated value ; the analyte concentration was less than the LOQ.
U	Analyte was not detected. The concentration is below the reported LOD.

Microbac

Lab Report #: L16061026
Lab Project #: 2551.096
Project Name: Longhorn Army Ammunition
Lab Contact: Michelle Taylor

2.3.2.2 QC Summary Data

Example 6020 Calculations
Perkin Elmer NexION 300X

1.0 Initial Calibration (ICAL) Parameters
The system performs linear regression from data consisting of a blank and three standards.

2.0 Calculating the concentration (C) of an element in water using data from prep log, run log, and quantitation report (note:the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (ug/L)	0.1
Vf = Final volume	100
Vi = Initial volume	40
D = Dilution factor as a multiplier (10X = 10)	1
Cx = Concentration of element in (ug/L)	0.25

3.0 Calculating the concentration (C) of an element in soil using data from prep log, run log, and quantitation report (note: the data system performs this calculation automatically when correction factors have been entered):

$$Cx = Cs \times \frac{Vf}{Vi} \times D$$

Where:	**Example:**
Cs = Concentration computed by the data system (ug/L)	0.1
Vf = Final volume	200
Vi = Initial volume	0.5
D = Dilution factor as a multiplier (10X = 10)	1
Cx = Concentration of element in (ug/kg)	40

4.0 Adjusting the concentration to dry weight:

$$Cdry = \frac{Cx \times 100}{Px}$$

Where:	**Example:**
Cx = Concentration calculated as received (wet basis)	40
Px = Percent solids of sample (%wt)	80
$Cdry$ = Concentration calculated as dry weight (ug/kg)	50

50 ug/kg = 0.050 mg/kg

Perkin Elmer NexION ICP/MS

STANDARDS KEY

QC Std 1 - ICV
QC Std 2 - ICB
QC Std 3 - LLICV
QC Std 4 - ICSA
QC Std 5 - ICSAB
QC Std 6 - CCV
QC Std 7 - CCB
QC Std 8 - LLCCV

Calibration Solutions

Analyte	Stock Conc. (mg/L)	S1 (mg/L)	S2 (mg/L)	S3 (mg/L)	S4 (mg/L)
Al	10	0	0.00005	0.05	0.1
Sb	10	0	0.00005	0.05	0.1
As	10	0	0.00005	0.05	0.1
Ba	10	0	0.00005	0.05	0.1
Be	10	0	0.00005	0.05	0.1
Ca	1000	0	0.005	5	10
Cd	10	0	0.0005	0.05	0.1
Cr	10	0	0.0005	0.05	0.1
Co	10	0	0.0005	0.05	0.1
Cu	10	0	0.0005	0.05	0.1
Fe	1000	0	0.005	5	10
Pb	10	0	0.00005	0.05	0.1
Mg	1000	0	0.005	5	10
Mn	10	0	0.00005	0.05	0.1
Ni	10	0	0.00005	0.05	0.1
K	1000	0	0.005	5	10
Se	10	0	0.00005	0.05	0.1
Ag	10	0	0.00005	0.05	0.1
Na	1000	0	0.005	5	10
Tl	10	0	0.00005	0.05	0.1
V	10	0	0.00005	0.05	0.1
U	1000	0	0.00005	0.05	0.1
Zn	10	0	0.00005	0.05	0.1

Microbac Laboratories Inc.
Microwave Digestion Log

Workgroup: WG573548
Analyst: VC
Spike Analyst: VC
Run Date: 06/21/2016 11:18
Method: 3015
Balance: BAL016
Instrument: MW-4
Instrument Start: 06/21/2016 11:33

SOP: ME407 Revison 19
Spike Solution: STD75998
Spike Witness: ERP
40 & 50 ML. DIGESTION TU COA18987
HNO3 Lot #: COA18905
MS Filters- fisher-Lot#R RGT35621

	SAMPLE #	Type	Matrix	Initial Amount	Final Volume	Initial Vessel Wt	Final Vessel Wt	Spike Amount	Due Date
1	WG573548-02	BLANK	1	20 mL	50 mL	181.353 g	181.33 g		
2	WG573548-03	LCS	1	20 mL	50 mL	182.64 g	182.591 g	.25 mL	
3	L16060986-01	SAMP	1	20 mL	50 mL	183.712 g	183.69 g		06/27/16
4	L16060986-03	SAMP	1	20 mL	50 mL	184.065 g	184.038 g		06/27/16
5	L16060986-05	SAMP	1	20 mL	50 mL	184.546 g	184.518 g		06/27/16
6	L16060986-07	SAMP	1	20 mL	50 mL	183.279 g	183.251 g		06/27/16
7	L16060986-09	SAMP	1	20 mL	50 mL	180.961 g	180.921 g		06/27/16
8	L16060986-11	SAMP	1	20 mL	50 mL	183.322 g	183.289 g		06/27/16
9	L16060986-13	SAMP	1	20 mL	50 mL	182.911 g	182.887 g		06/27/16
10	L16060986-15	SAMP	1	20 mL	50 mL	180.852 g	180.836 g		06/27/16
11	L16060986-17	SAMP	1	20 mL	50 mL	184.362 g	184.347 g		06/27/16
12	L16060986-19	SAMP	1	20 mL	50 mL	181.353 g	181.343 g		06/27/16
13	WG573548-01	REF	1	20 mL	50 mL	182.733 g	182.717 g		
14	L16060986-23	RS03	1	20 mL	50 mL	182.733 g	182.717 g		06/27/16
15	L16060986-25	SAMP	1	20 mL	50 mL	183.941 g	183.923 g		06/27/16
16	L16060986-27	SAMP	1	20 mL	50 mL	180.794 g	180.793 g		06/27/16
17	L16060986-29	SAMP	1	20 mL	50 mL	182.177 g	182.161 g		06/27/16
18	WG573548-04	MS	1	20 mL	50 mL	182.271 g	182.257 g	.25 mL	
19	L16060986-35	MS03	1	20 mL	50 mL	182.271 g	182.257 g	.25 mL	06/27/16
20	WG573548-05	MSD	1	20 mL	50 mL	184.375 g	184.35 g	.25 mL	
21	L16060986-37	SD03	1	20 mL	50 mL	184.375 g	184.35 g	.25 mL	06/27/16
22	L16060986-39	SAMP	1	20 mL	50 mL	181.551 g	181.533 g		06/27/16
23	L16060986-41	SAMP	1	20 mL	50 mL	184.38 g	184.379 g		06/27/16
24	L16061026-02	SAMP	1	20 mL	50 mL	182.267 g	182.259 g		06/28/16
25	L16061026-04	SAMP	1	20 mL	50 mL	182.882 g	182.893 g		06/28/16
26	L16061026-06	SAMP	1	20 mL	50 mL	180.796 g	180.773 g		06/28/16
27	L16061026-07	SAMP	1	20 mL	50 mL	181.26 g	181.246 g		06/28/16

L16060986-01	FILTERED DIGESTATE
L16060986-09	FILTERED DIGESTATE
L16060986-13	FILTERED DIGESTATE
L16060986-15	FILTERED DIGESTATE
L16060986-25	FILTERED DIGESTATE
L16060986-27	FILTERED DIGESTATE
L16060986-29	FILTERED DIGESTATE

MW_DIG - Modified 09/30/2009
PDF ID: 4816619
Report generated: 06/21/2016 12:51

Microbac Laboratories Inc.
Microwave Digestion Log

Analyst: Vicki Collins

SOP:
Spike Solution:
Reviewer: Sam Potter
Spike Witness:

Method:
Balance:
Instrument:
Instrument Start:

MW_DIG - Modified 09/30/2009
PDF ID: 4816619
Report generated: 06/21/2016 12:51

Microbac

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	NI.062916.112234	Blank	Blank		1		06/29/16 11:22
2	NI.062916.112545	WG574676-01	Calibration Point		1		06/29/16 11:25
3	NI.062916.112857	WG574676-02	Calibration Point		1		06/29/16 11:28
4	NI.062916.113208	WG574676-03	Calibration Point		1		06/29/16 11:32
5	NI.062916.113520	WG574676-04	Calibration Point		1		06/29/16 11:35
6	NI.062916.113833	WG574676-05	Initial Calibration Verification		1		06/29/16 11:38
7	NI.062916.114146	WG574676-06	Initial Calib Blank		1		06/29/16 11:41
8	NI.062916.114653	WG574676-07	Low Level Initial Calibration V		1		06/29/16 11:46
9	NI.062916.115053	WG574676-08	Interference Check		1		06/29/16 11:50
10	NI.062916.115447	WG574676-09	Interference Check		1		06/29/16 11:54
11	NI.062916.120010	WG574676-10	CCV		1		06/29/16 12:00
12	NI.062916.120734	WG574676-11	CCB		1		06/29/16 12:07
13	NI.062916.121215	L16060661-06	US23SS02502A	.253/100	1		06/29/16 12:12
14	NI.062916.121526	L16060661-07	US23SS02505A	.257/100	1		06/29/16 12:15
15	NI.062916.121838	L16060661-08	US23SS02602A	.257/100	1		06/29/16 12:18
16	NI.062916.122149	L16060661-09	US23SS02605A	.25/100	1		06/29/16 12:21
17	NI.062916.122501	L16060822-03	18CPTMW04SW-061416	20/50	10		06/29/16 12:25
18	NI.062916.122812	L16060823-02	126F-061416	20/50	50		06/29/16 12:28
19	NI.062916.123124	L16060823-03	C08-061416	20/50	50		06/29/16 12:31
20	NI.062916.123435	L16060823-04	C03-061416	20/50	50		06/29/16 12:34
21	NI.062916.123748	WG574676-12	CCV		1		06/29/16 12:37
22	NI.062916.124059	WG574676-13	CCB		1		06/29/16 12:40
23	NI.062916.124412	WG574676-14	Low Level Continuing Calibra		1		06/29/16 12:44
24	NI.062916.125354	WG573548-02	Method/Prep Blank	20/50	1		06/29/16 12:53
25	NI.062916.125705	WG573548-03	Laboratory Control S	20/50	1		06/29/16 12:57
26	NI.062916.130017	WG573548-01	Reference Sample		1	L16060986-23	06/29/16 13:00
27	NI.062916.130328	WG573548-04	Matrix Spike	20/50	1	L16060986-23	06/29/16 13:03
28	NI.062916.130639	WG573548-05	Matrix Spike Duplica	20/50	1	L16060986-23	06/29/16 13:06
29	NI.062916.130951	L16060986-01	MW-07-04	20/50	1		06/29/16 13:09
30	NI.062916.131302	L16060986-03	MW-01-10	20/50	1		06/29/16 13:13
31	NI.062916.131614	WG574572-01	Post Digestion Spike		1	L16060986-03	06/29/16 13:16
32	NI.062916.131925	WG574572-02	Serial Dilution		5	L16060986-03	06/29/16 13:19
33	NI.062916.132237	WG574572-02	Serial Dilution		25	L16060986-03	06/29/16 13:22
34	NI.062916.132550	WG574676-15	CCV		1		06/29/16 13:25

Page: 1 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	NI.062916.132901	WG574676-16	CCB		1		06/29/16 13:29
36	NI.062916.133214	L16060986-05	MW-03-12	20/50	1		06/29/16 13:32
37	NI.062916.133526	L16060986-07	MW-01-07	20/50	1		06/29/16 13:35
38	NI.062916.133837	L16060986-09	MW-03-11	20/50	1		06/29/16 13:38
39	NI.062916.134150	L16060986-11	MW-01-01	20/50	1		06/29/16 13:41
40	NI.062916.134502	L16060986-13	MW-03-09	20/50	1		06/29/16 13:45
41	NI.062916.134813	L16060986-15	MW-01-08	20/50	1		06/29/16 13:48
42	NI.062916.135125	L16060986-17	MW-03-06	20/50	1		06/29/16 13:51
43	NI.062916.135436	L16060986-19	MW-03-05	20/50	1		06/29/16 13:54
44	NI.062916.135748	L16060986-25	MW-01-02	20/50	1		06/29/16 13:57
45	NI.062916.140059	L16060986-27	DUP 1	20/50	1		06/29/16 14:00
46	NI.062916.140412	WG574676-17	CCV		1		06/29/16 14:04
47	NI.062916.140723	WG574676-18	CCB		1		06/29/16 14:07
48	NI.062916.141036	L16060986-29	DUP 2	20/50	1		06/29/16 14:10
49	NI.062916.141347	L16060986-39	EQUIP BLANK 1	20/50	1		06/29/16 14:13
50	NI.062916.141659	L16060986-41	EQUIP BLANK 2	20/50	1		06/29/16 14:16
51	NI.062916.145120	L16061026-02	18WW16F-061616	20/50	1		06/29/16 14:51
52	NI.062916.145431	L16061026-04	129-061616	20/50	1		06/29/16 14:54
53	NI.062916.145742	L16061026-06	120F-061616	20/50	1		06/29/16 14:57
54	NI.062916.150054	L16061026-07	123-061616	20/50	1		06/29/16 15:00
55	NI.062916.150407	WG574676-19	CCV		1		06/29/16 15:04
56	NI.062916.150718	WG574676-20	CCB		1		06/29/16 15:07
57	NI.062916.151032	WG573099-02	Method/Prep Blank	.25/100	1		06/29/16 15:10
58	NI.062916.151343	WG573099-03	Laboratory Control S	.25/100	1		06/29/16 15:13
59	NI.062916.151655	WG573099-01	Reference Sample		10	L16060888-03	06/29/16 15:16
60	NI.062916.152006	WG573099-04	Matrix Spike	.255/100	10	L16060888-03	06/29/16 15:20
61	NI.062916.152318	WG573099-05	Matrix Spike Duplica	.254/100	10	L16060888-03	06/29/16 15:23
62	NI.062916.152629	L16060883-01	WC01	.252/100	1		06/29/16 15:26
63	NI.062916.152941	L16060912-01	WH8011	.25/100	1		06/29/16 15:29
64	NI.062916.153252	WG573717-03	Post Digestion Spike		1	L16060912-01	06/29/16 15:32
65	NI.062916.153604	WG573717-04	Serial Dilution		5	L16060912-01	06/29/16 15:36
66	NI.062916.153915	WG573717-04	Serial Dilution		25	L16060912-01	06/29/16 15:39
67	NI.062916.154228	WG574676-21	CCV		1		06/29/16 15:42
68	NI.062916.154540	WG574676-22	CCB		1		06/29/16 15:45

Page: 2 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
69	NI.062916.154853	WG574001-04	Method/Prep Blank	20/50	1		06/29/16 15:48
70	NI.062916.155204	WG574001-05	Laboratory Control S	20/50	1		06/29/16 15:52
71	NI.062916.155516	WG574001-01	Reference Sample		1	L16061247-02	06/29/16 15:55
72	NI.062916.155827	WG574001-06	Duplicate	20/50	1	L16061247-02	06/29/16 15:58
73	NI.062916.160139	WG574001-02	Reference Sample		1	L16061265-01	06/29/16 16:01
74	NI.062916.160450	WG574001-07	Matrix Spike	20/50	1	L16061265-01	06/29/16 16:04
75	NI.062916.160802	WG574001-08	Matrix Spike Duplica	20/50	1	L16061265-01	06/29/16 16:08
76	NI.062916.161113	L16061247-01	WMB09	20/50	1		06/29/16 16:11
77	NI.062916.161425	WG574688-01	Post Digestion Spike		1	L16061247-01	06/29/16 16:14
78	NI.062916.161736	WG574688-02	Serial Dilution		5	L16061247-01	06/29/16 16:17
79	NI.062916.162049	WG574676-23	CCV		1		06/29/16 16:20
80	NI.062916.162401	WG574676-24	CCB		1		06/29/16 16:24
81	NI.062916.162713	WG574688-02	Serial Dilution		25	L16061247-01	06/29/16 16:27
82	NI.062916.163025	L16061265-04	SS007MWOB16-007	20/50	1		06/29/16 16:30
83	NI.062916.163336	L16061265-05	SS007MWOB21-007	20/50	1		06/29/16 16:33
84	NI.062916.163648	L16061265-06	FQC-RB06211601-CC	20/50	1		06/29/16 16:36
85	NI.062916.163959	L16061266-01	ST008MW0805-007	20/50	1		06/29/16 16:39
86	NI.062916.164310	L16061269-01	FT002MW0204-007	20/50	1		06/29/16 16:43
87	NI.062916.164622	L16061272-01	SS007MW0B18-007	20/50	1		06/29/16 16:46
88	NI.062916.164933	L16061272-02	SS007MWN752-007	20/50	1		06/29/16 16:49
89	NI.062916.165245	L16061279-01	XU936TW007A	20/50	1		06/29/16 16:52
90	NI.062916.165556	L16061279-01	XU936TW007A	20/50	50		06/29/16 16:55
91	NI.062916.165909	WG574676-25	CCV		1		06/29/16 16:59
92	NI.062916.170220	WG574676-26	CCB		1		06/29/16 17:02
93	NI.062916.170533	WG574001-03	Reference Sample		1	L16061279-02	06/29/16 17:05
94	NI.062916.170845	WG574001-03	Reference Sample		50	L16061279-02	06/29/16 17:08
95	NI.062916.171156	WG574001-09	Matrix Spike	20/50	1	L16061279-02	06/29/16 17:11
96	NI.062916.171508	WG574001-09	Matrix Spike	20/50	50	L16061279-02	06/29/16 17:15
97	NI.062916.171819	WG574001-10	Matrix Spike Duplica	20/50	1	L16061279-02	06/29/16 17:18
98	NI.062916.172131	WG574001-10	Matrix Spike Duplica	20/50	50	L16061279-02	06/29/16 17:21
99	NI.062916.172442	L16061279-05	XU936TW019A	20/50	1		06/29/16 17:24
100	NI.062916.172754	L16061279-05	XU936TW019A	20/50	50		06/29/16 17:27
101	NI.062916.173105	L16061281-01	J6F0803-01	20/50	1		06/29/16 17:31
102	NI.062916.173419	WG574676-27	CCV		1		06/29/16 17:34

Page: 3 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP
Analyst1: BKT Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705
ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607
CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345
Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
103	NI.062916.173730	WG574676-28	CCB		1		06/29/16 17:37
104	NI.062916.174043	L16061294-03	EB01	20/50	1		06/29/16 17:40
105	NI.062916.174355	L16061294-03	EB01	20/50	50		06/29/16 17:43
106	NI.062916.174706	L16061296-01	OUTFALL 002/COMP	20/50	1		06/29/16 17:47
107	NI.062916.175019	WG574676-29	Interference Check		1		06/29/16 17:50
108	NI.062916.175331	WG574676-30	Interference Check		1		06/29/16 17:53
109	NI.062916.175644	WG574676-31	CCV		1		06/29/16 17:56
110	NI.062916.175955	WG574676-32	CCB		1		06/29/16 17:59
111	NI.062916.180308	WG574676-33	Low Level Continuing Calibra		1		06/29/16 18:03
112	NI.062916.180621	WG574192-02	Method/Prep Blank	.25/100	1		06/29/16 18:06
113	NI.062916.180932	WG574192-03	Laboratory Control S	.25/100	1		06/29/16 18:09
114	NI.062916.181244	WG574192-01	Reference Sample		1	L16061206-01	06/29/16 18:12
115	NI.062916.181555	L16061206-02	LF003SD0311-007 DUP	.251/100	1		06/29/16 18:15
116	NI.062916.181907	WG574192-04	Matrix Spike	.25/100	1	L16061206-01	06/29/16 18:19
117	NI.062916.182218	WG574192-05	Matrix Spike Duplica	.253/100	1	L16061206-01	06/29/16 18:22
118	NI.062916.182530	L16061206-05	LF003SD0312-007	.25/100	1		06/29/16 18:25
119	NI.062916.182841	WG574690-01	Post Digestion Spike		1	L16061206-05	06/29/16 18:28
120	NI.062916.183153	WG574690-02	Serial Dilution		5	L16061206-05	06/29/16 18:31
121	NI.062916.183504	WG574690-02	Serial Dilution		25	L16061206-05	06/29/16 18:35
122	NI.062916.183818	WG574676-34	CCV		1		06/29/16 18:38
123	NI.062916.184129	WG574676-35	CCB		1		06/29/16 18:41
124	NI.062916.184442	L16061207-01	SS007SD0709-007	.251/100	1		06/29/16 18:44
125	NI.062916.184754	L16061207-02	SS007SD0710-007	.25/100	1		06/29/16 18:47
126	NI.062916.185105	L16061208-01	DP005SD0513-007	.252/100	1		06/29/16 18:51
127	NI.062916.185417	L16061208-02	DP005SD0514-007	.251/100	1		06/29/16 18:54
128	NI.062916.185728	L16061291-01	V6F0270-01	.253/100	1		06/29/16 18:57
129	NI.062916.190040	L16061381-01	XU936DS00500A	.254/100	1		06/29/16 19:00
130	NI.062916.190351	L16061381-01	XU936DS00500A	.254/100	50		06/29/16 19:03
131	NI.062916.190703	L16061381-02	XU936DS00505A	.258/100	1		06/29/16 19:07
132	NI.062916.191014	L16061381-02	XU936DS00505A	.258/100	50		06/29/16 19:10
133	NI.062916.191327	WG574676-36	CCV		1		06/29/16 19:13
134	NI.062916.191639	WG574676-37	CCB		1		06/29/16 19:16
135	NI.062916.191952	L16061381-03	XU936DS00300A	.252/100	1		06/29/16 19:19
136	NI.062916.192303	L16061381-03	XU936DS00300A	.252/100	50		06/29/16 19:23

Page: 4 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 062916A.REP

Analyst1: BKT Analyst2: N/A

Method: 6020/6020A/200.8 SOP: ME700A Rev: 2

Maintenance Log ID:

Calibration Std: STD76147 ICV Std: STD76279 Post Spike: STD73705

ICSA: STD76453 ICSAB: STD76213 Int. Std: RGT36607

CCV: STD75971 LLCCV: STD76223 Tuning Sol : STD76345

Stannous : Hydroxylamine :

Workgroups: 573722,574570,574572,573717,574688,574690,573790

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
137	NI.062916.192615	L16061381-04	XU936DS00305A	.25/100	1		06/29/16 19:26
138	NI.062916.192926	L16061381-04	XU936DS00305A	.25/100	50		06/29/16 19:29
139	NI.062916.193237	L16061381-05	XU936DS00305C	.253/100	1		06/29/16 19:32
140	NI.062916.193550	L16061381-05	XU936DS00305C	.253/100	50		06/29/16 19:35
141	NI.062916.193901	L16061381-07	XU936DS00400A	.252/100	1		06/29/16 19:39
142	NI.062916.194213	L16061381-07	XU936DS00400A	.252/100	50		06/29/16 19:42
143	NI.062916.194524	L16061384-02	BFB-16-019	.253/100	1		06/29/16 19:45
144	NI.062916.194835	L16061384-04	BFB-16-020	.253/100	1		06/29/16 19:48
145	NI.062916.195149	WG574676-38	CCV		1		06/29/16 19:51
146	NI.062916.195500	WG574676-39	CCB		1		06/29/16 19:55
147	NI.062916.195813	L16061384-06	BFB-16-021	.253/100	1		06/29/16 19:58
148	NI.062916.200124	L16061384-08	BFB-16-022	.251/100	1		06/29/16 20:01
149	NI.062916.200436	L16061384-10	BFB-16-023	.254/100	1		06/29/16 20:04
150	NI.062916.200747	L16061384-12	BFB-16-024	.254/100	1		06/29/16 20:07
151	NI.062916.201101	WG574676-40	CCV		1		06/29/16 20:11
152	NI.062916.201412	WG574676-41	CCB		1		06/29/16 20:14
153	NI.062916.201726	WG573291-03	Method/Prep Blank	20/50	1		06/29/16 20:17
154	NI.062916.202037	WG573291-04	Laboratory Control S	20/50	1		06/29/16 20:20
155	NI.062916.202348	WG573291-01	Reference Sample		1	L16060974-01	06/29/16 20:23
156	NI.062916.202700	WG573291-05	Matrix Spike	20/50	1	L16060974-01	06/29/16 20:27
157	NI.062916.203011	WG573291-06	Matrix Spike Duplica	20/50	1	L16060974-01	06/29/16 20:30
158	NI.062916.203323	WG573790-03	Post Digestion Spike		1	L16060974-01	06/29/16 20:33
159	NI.062916.203634	WG573790-04	Serial Dilution		5	L16060974-01	06/29/16 20:36
160	NI.062916.203946	WG573790-04	Serial Dilution		25	L16060974-01	06/29/16 20:39
161	NI.062916.204259	WG574676-42	CCV		1		06/29/16 20:42
162	NI.062916.204611	WG574676-43	CCB		1		06/29/16 20:46

Page: 5 Approved: July 05, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070516A.REP
Analyst1: JYH Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD76567
ICSA: STD76453 ICSAB: STD76573 Int. Std: RGT36607
CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989
Stannous : Hydroxylamine :

Workgroups: 575049,575057,575052,575054,574236,574690,574572,575217

Comments: Additional Workgroup: 574822

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	NI.070516.092352	Blank	Blank		1		07/05/16 09:23
2	NI.070516.092704	WG575179-01	Calibration Point		1		07/05/16 09:27
3	NI.070516.093015	WG575179-02	Calibration Point		1		07/05/16 09:30
4	NI.070516.093327	WG575179-03	Calibration Point		1		07/05/16 09:33
5	NI.070516.093638	WG575179-04	Calibration Point		1		07/05/16 09:36
6	NI.070516.093952	WG575179-05	ICV		1		07/05/16 09:39
7	NI.070516.094305	WG575179-06	ICB		1		07/05/16 09:43
8	NI.070516.094618	WG575179-07	LLICV		1		07/05/16 09:46
9	NI.070516.095016	WG575179-08	Initial Calibration Verification		1		07/05/16 09:50
10	NI.070516.095329	WG575179-09	Initial Calib Blank		1		07/05/16 09:53
11	NI.070516.095642	WG575179-10	Low Level Initial Calibration V		1		07/05/16 09:56
12	NI.070516.095958	WG575179-11	Interference Check		1		07/05/16 09:59
13	NI.070516.100317	WG575179-12	Interference Check		1		07/05/16 10:03
14	NI.070516.100630	WG575179-13	CCV		1		07/05/16 10:06
15	NI.070516.100941	WG575179-14	CCB		1		07/05/16 10:09
16	NI.070516.101530	WG574663-02	Method/Prep Blank	.25/100	1		07/05/16 10:15
17	NI.070516.101841	WG574663-03	Laboratory Control S	.25/100	1		07/05/16 10:18
18	NI.070516.102152	WG574663-01	Reference Sample		1	L16061451-06	07/05/16 10:21
19	NI.070516.102504	WG574663-04	Matrix Spike	.252/100	1	L16061451-06	07/05/16 10:25
20	NI.070516.102815	WG574663-05	Matrix Spike Duplica	.253/100	1	L16061451-06	07/05/16 10:28
21	NI.070516.103127	L16061451-01	GS165SS062016S-ES	.254/100	1		07/05/16 10:31
22	NI.070516.103438	L16061451-02	GS159SS062116G-ES	.251/100	1		07/05/16 10:34
23	NI.070516.103750	WG575049-01	Post Digestion Spike		1	L16061451-02	07/05/16 10:37
24	NI.070516.104101	WG575049-02	Serial Dilution		5	L16061451-02	07/05/16 10:41
25	NI.070516.104413	WG575049-02	Serial Dilution		25	L16061451-02	07/05/16 10:44
26	NI.070516.104726	WG575179-15	CCV		1		07/05/16 10:47
27	NI.070516.105037	WG575179-16	CCB		1		07/05/16 10:50
28	NI.070516.105350	L16061451-03	GS164SS062116S-ES	.252/100	1		07/05/16 10:53
29	NI.070516.105702	L16061451-04	GS164SS062116P-ES	.257/100	1		07/05/16 10:57
30	NI.070516.110013	L16061451-05	GS154SS062316X-ES	.252/100	1		07/05/16 11:00
31	NI.070516.110324	L16061451-07	GS153SS0622316X-ES	.256/100	1		07/05/16 11:03
32	NI.070516.110636	L16061451-08	GS153SS062316X-ESD	.256/100	1		07/05/16 11:06
33	NI.070516.110948	L16061445-03	TS935DS00800A	.255/100	1		07/05/16 11:09
34	NI.070516.111259	L16061445-07	TS935DS00900A		1		07/05/16 11:12

Page: 1 Approved: July 06, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070516A.REP
Analyst1: JYH Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD76567
ICSA: STD76453 ICSAB: STD76573 Int. Std: RGT36607
CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989
Stannous : Hydroxylamine :

Workgroups: 575049,575057,575052,575054,574236,574690,574572,575217

Comments: Additional Workgroup: 574822

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	NI.070516.111611	L16061445-08	TS935DS00905A	.255/100	1		07/05/16 11:16
36	NI.070516.111922	L16061445-09	TS935DS00905C	.256/100	1		07/05/16 11:19
37	NI.070516.112234	L16061445-10	TS935DS01000A	.256/100	1		07/05/16 11:22
38	NI.070516.112546	WG575179-17	CCV		1		07/05/16 11:25
39	NI.070516.112857	WG575179-18	CCB		1		07/05/16 11:28
40	NI.070516.113210	L16061445-11	TS935DS01005A	.256/100	1		07/05/16 11:32
41	NI.070516.113521	L16061445-12	TS935DS01205A	.254/100	1		07/05/16 11:35
42	NI.070516.113833	L16061451-01	GS165SS062016S-ES	.251/100	10		07/05/16 11:38
43	NI.070516.114144	L16061451-02	GS159SS062116G-ES	.25/100	10		07/05/16 11:41
44	NI.070516.114455	WG575057-01	Post Digestion Spike		10	L16061451-02	07/05/16 11:44
45	NI.070516.114807	WG575057-02	Serial Dilution		50	L16061451-02	07/05/16 11:48
46	NI.070516.115118	L16061451-03	GS164SS062116S-ES	.251/100	10		07/05/16 11:51
47	NI.070516.115430	L16061451-04	GS164SS062116P-ES	.253/100	10		07/05/16 11:54
48	NI.070516.115741	L16061451-05	GS154SS062316X-ES		10		07/05/16 11:57
49	NI.070516.120052	WG574521-02	Reference Sample		10	L16061451-06	07/05/16 12:00
50	NI.070516.120406	WG575179-19	CCV		1		07/05/16 12:04
51	NI.070516.120717	WG575179-20	CCB		1		07/05/16 12:07
52	NI.070516.121030	L16061451-07	GS153SS0622316X-ES	.253/100	10		07/05/16 12:10
53	NI.070516.121342	L16061451-08	GS153SS062316X-ESD	.256/100	10		07/05/16 12:13
54	NI.070516.121654	WG574521-07	Matrix Spike	.258/100	10	L16061451-06	07/05/16 12:16
55	NI.070516.122005	WG574521-08	Matrix Spike Duplica	.256/100	10	L16061451-06	07/05/16 12:20
56	NI.070516.122318	WG575179-21	CCV		1		07/05/16 12:23
57	NI.070516.122629	WG575179-22	CCB		1		07/05/16 12:26
58	NI.070516.122942	WG574664-02	Method/Prep Blank	20/50	1		07/05/16 12:29
59	NI.070516.123254	WG574664-03	Laboratory Control S	20/50	1		07/05/16 12:32
60	NI.070516.123605	WG574664-01	Reference Sample		1	L16061548-01	07/05/16 12:36
61	NI.070516.123917	WG574664-04	Duplicate	20/50	1	L16061548-01	07/05/16 12:39
62	NI.070516.124228	WG574664-05	Matrix Spike	20/50	1	L16061548-01	07/05/16 12:42
63	NI.070516.124540	WG574664-06	Matrix Spike Duplica	20/50	1	L16061548-01	07/05/16 12:45
64	NI.070516.124851	L16061544-01	S6F0335-01	20/50	1		07/05/16 12:48
65	NI.070516.125203	WG575052-01	Post Digestion Spike		1	L16061544-01	07/05/16 12:52
66	NI.070516.125514	WG575052-02	Serial Dilution		5	L16061544-01	07/05/16 12:55
67	NI.070516.125826	WG575052-02	Serial Dilution		25	L16061544-01	07/05/16 12:58
68	NI.070516.130139	WG575179-23	CCV		1		07/05/16 13:01

Page: 2 Approved: July 06, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070516A.REP
Analyst1: JYH Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD76567
ICSA: STD76453 ICSAB: STD76573 Int. Std: RGT36607
CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989
Stannous : Hydroxylamine :

Workgroups: 575049,575057,575052,575054,574236,574690,574572,575217

Comments: Additional Workgroup: 574822

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
69	NI.070516.130450	WG575179-24	CCB		1		07/05/16 13:04
70	NI.070516.130802	L16061546-01	S6F0336-01	20/50	1		07/05/16 13:08
71	NI.070516.131114	L16061547-01	S6F0337-01	20/50	1		07/05/16 13:11
72	NI.070516.131425	L16061313-02	MW-5	20/50	1		07/05/16 13:14
73	NI.070516.131737	L16061313-03	MW-5	20/50	1		07/05/16 13:17
74	NI.070516.132050	WG575179-25	CCV		1		07/05/16 13:20
75	NI.070516.132401	WG575179-26	CCB		1		07/05/16 13:24
76	NI.070516.132714	WG574733-02	Method/Prep Blank	20/50	1		07/05/16 13:27
77	NI.070516.133026	WG574733-03	Laboratory Control S	20/50	1		07/05/16 13:30
78	NI.070516.133337	WG574733-01	Reference Sample		1	L16061556-06	07/05/16 13:33
79	NI.070516.133649	WG574733-04	Duplicate	20/50	1	L16061556-06	07/05/16 13:36
80	NI.070516.134000	WG574733-05	Matrix Spike	20/50	1	L16061556-06	07/05/16 13:40
81	NI.070516.134312	WG574733-06	Matrix Spike Duplica	20/50	1	L16061556-06	07/05/16 13:43
82	NI.070516.134623	L16061644-01	OUTFALL 002/COMP	20/50	1		07/05/16 13:46
83	NI.070516.134935	WG575054-01	Post Digestion Spike		1	L16061644-01	07/05/16 13:49
84	NI.070516.135246	WG575054-02	Serial Dilution		5	L16061644-01	07/05/16 13:52
85	NI.070516.135719	L16061445-07	TS935DS00900A	.25/100	10		07/05/16 13:57
86	NI.070516.140031	WG575179-27	CCV		1		07/05/16 14:00
87	NI.070516.140343	WG575179-28	CCB		1		07/05/16 14:03
88	NI.070516.140655	L16061559-01	LH18/24-SP650-6374-GRAB	20/50	1		07/05/16 14:06
89	NI.070516.141021	L16061559-02	LH18/24-SP650-6374-COMP	20/50	1		07/05/16 14:10
90	NI.070516.141332	L16061581-03	OUTFALL 203		1		07/05/16 14:13
91	NI.070516.141644	WG574664-01	Reference Sample		2	L16061548-01	07/05/16 14:16
92	NI.070516.141955	WG574664-04	Duplicate	20/50	2	L16061548-01	07/05/16 14:19
93	NI.070516.142314	WG574664-05	Matrix Spike	20/50	2	L16061548-01	07/05/16 14:23
94	NI.070516.142626	WG574664-06	Matrix Spike Duplica	20/50	2	L16061548-01	07/05/16 14:26
95	NI.070516.142939	WG575179-29	CCV		1		07/05/16 14:29
96	NI.070516.143251	WG575179-30	CCB		1		07/05/16 14:32
97	NI.070516.143603	WG575179-31	Low Level Continuing Calibra		1		07/05/16 14:36
98	NI.070516.143943	WG573653-03	Method/Prep Blank	20/50	1		07/05/16 14:39
99	NI.070516.144254	WG573653-04	Laboratory Control S	20/50	1		07/05/16 14:42
100	NI.070516.144606	WG573653-01	Reference Sample		1	L16060764-04	07/05/16 14:46
101	NI.070516.144917	WG573653-05	Matrix Spike	20/50	1	L16060764-04	07/05/16 14:49
102	NI.070516.145229	WG573653-06	Matrix Spike Duplica	20/50	1	L16060764-04	07/05/16 14:52

Page: 3 Approved: July 06, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070516A.REP
Analyst1: JYH Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD76567
ICSA: STD76453 ICSAB: STD76573 Int. Std: RGT36607
CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989
Stannous : Hydroxylamine :

Workgroups: 575049,575057,575052,575054,574236,574690,574572,575217

Comments: Additional Workgroup: 574822

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
103	NI.070516.145558	L16060840-01	MW-24-004R	20/50	1		07/05/16 14:55
104	NI.070516.145909	L16060840-02	MW-24-003R	20/50	1		07/05/16 14:59
105	NI.070516.150221	WG574236-03	Post Digestion Spike		1	L16060840-02	07/05/16 15:02
106	NI.070516.150532	WG574236-04	Serial Dilution		5	L16060840-02	07/05/16 15:05
107	NI.070516.150843	L16061581-03	OUTFALL 203	20/50	1000		07/05/16 15:08
108	NI.070516.151157	WG575179-32	CCV		1		07/05/16 15:11
109	NI.070516.151517	WG575179-33	CCB		1		07/05/16 15:15
110	NI.070516.151839	L16060840-03	FR-061416-24	20/50	1		07/05/16 15:18
111	NI.070516.152150	L16060840-04	MW-24-002R	20/50	1		07/05/16 15:21
112	NI.070516.152502	L16060840-05	MW-24-005	20/50	1		07/05/16 15:25
113	NI.070516.152813	L16060840-06	EB-LOW FLOW TUBING	20/50	1		07/05/16 15:28
114	NI.070516.153125	L16061291-01	V6F0270-01	.253/100	50		07/05/16 15:31
115	NI.070516.153436	L16061533-04	EB05	20/50	1		07/05/16 15:34
116	NI.070516.153748	L16061533-05	EB04	20/50	1		07/05/16 15:37
117	NI.070516.154100	WG575179-34	CCV		1		07/05/16 15:41
118	NI.070516.154411	WG575179-35	CCB		1		07/05/16 15:44
119	NI.070516.154936	WG573548-02	Method/Prep Blank	20/50	1		07/05/16 15:49
120	NI.070516.155247	WG573548-03	Laboratory Control S	20/50	1		07/05/16 15:52
121	NI.070516.155559	WG573548-01	Reference Sample		1	L16060986-23	07/05/16 15:55
122	NI.070516.155910	WG573548-04	Matrix Spike	20/50	1	L16060986-23	07/05/16 15:59
123	NI.070516.160221	WG573548-05	Matrix Spike Duplica	20/50	1	L16060986-23	07/05/16 16:02
124	NI.070516.160533	L16061026-02	18WW16F-061616	20/50	20		07/05/16 16:05
125	NI.070516.160845	L16061026-02	18WW16F-061616	20/50	200		07/05/16 16:08
126	NI.070516.161156	L16061026-04	129-061616	20/50	1		07/05/16 16:11
127	NI.070516.161508	L16061026-04	129-061616	20/50	20		07/05/16 16:15
128	NI.070516.161820	WG575179-36	CCV		1		07/05/16 16:18
129	NI.070516.162132	WG575179-37	CCB		1		07/05/16 16:21
130	NI.070516.163034	L16061026-06	120F-061616	20/50	1		07/05/16 16:30
131	NI.070516.163440	L16061026-07	123-061616	20/50	1		07/05/16 16:34
132	NI.070516.163751	WG574572-03	Post Digestion Spike		1	L16061026-07	07/05/16 16:37
133	NI.070516.164103	WG574572-04	Serial Dilution		5	L16061026-07	07/05/16 16:41
134	NI.070516.164414	WG574572-04	Serial Dilution		25	L16061026-07	07/05/16 16:44
135	NI.070516.164727	WG575179-38	CCV		1		07/05/16 16:47
136	NI.070516.165039	WG575179-39	CCB		1		07/05/16 16:50

Page: 4 Approved: July 06, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070516A.REP
Analyst1: JYH Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD76567
ICSA: STD76453 ICSAB: STD76573 Int. Std: RGT36607
CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989
Stannous : Hydroxylamine :

Workgroups: 575049,575057,575052,575054,574236,574690,574572,575217

Comments: Additional Workgroup: 574822

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
137	NI.070516.165352	WG575144-02	Method/Prep Blank	.25/100	1		07/05/16 16:53
138	NI.070516.165703	WG575144-03	Laboratory Control S	.25/100	1		07/05/16 16:57
139	NI.070516.170015	WG575144-01	Reference Sample		1	L16070088-01	07/05/16 17:00
140	NI.070516.170326	WG575144-04	Matrix Spike	.255/100	1	L16070088-01	07/05/16 17:03
141	NI.070516.170638	WG575144-05	Matrix Spike Duplica	.252/100	1	L16070088-01	07/05/16 17:06
142	NI.070516.170949	WG575217-01	Post Digestion Spike		1	L16070088-01	07/05/16 17:09
143	NI.070516.171301	WG575217-02	Serial Dilution		5	L16070088-01	07/05/16 17:13
144	NI.070516.171612	WG575217-02	Serial Dilution		25	L16070088-01	07/05/16 17:16
145	NI.070516.171925	WG575179-40	CCV		1		07/05/16 17:19
146	NI.070516.172236	WG575179-41	CCB		1		07/05/16 17:22
147	NI.070516.172549	WG573662-03	Method/Prep Blank	20/50	1		07/05/16 17:25
148	NI.070516.172901	WG573662-04	Laboratory Control S	20/50	1		07/05/16 17:29
149	NI.070516.173212	L16060907-01	MW21-061516	20/50	50		07/05/16 17:32
150	NI.070516.173523	L16060907-02	MW21-061516	20/50	50		07/05/16 17:35
151	NI.070516.173835	L16060907-04	MW35-061516	20/50	1		07/05/16 17:38
152	NI.070516.174147	L16060907-05	MW35-061516	20/50	1		07/05/16 17:41
153	NI.070516.174458	L16060907-07	MW11S-061516	20/50	1		07/05/16 17:44
154	NI.070516.174810	WG574822-01	Post Digestion Spike		1	L16060907-07	07/05/16 17:48
155	NI.070516.175121	L16060907-07	MW11S-061516		5		07/05/16 17:51
156	NI.070516.175433	WG574822-02	Serial Dilution		25	L16060907-07	07/05/16 17:54
157	NI.070516.175746	WG575179-42	CCV		1		07/05/16 17:57
158	NI.070516.180057	WG575179-43	CCB		1		07/05/16 18:00
159	NI.070516.180410	L16060907-08	MW11S-061516		1		07/05/16 18:04
160	NI.070516.180721	L16060907-10	DUP-GW-061516		1		07/05/16 18:07
161	NI.070516.181033	L16060907-11	DUP-GW-061516		1		07/05/16 18:10
162	NI.070516.181344	L16060907-12	MW31-061516		50		07/05/16 18:13
163	NI.070516.181656	L16060907-13	MW31-061516		50		07/05/16 18:16
164	NI.070516.182008	WG573662-01	Reference Sample		1	L16060907-14	07/05/16 18:20
165	NI.070516.182319	WG573662-02	Reference Sample		1	L16060907-15	07/05/16 18:23
166	NI.070516.182631	WG573662-05	Matrix Spike	20/50	1	L16060907-14	07/05/16 18:26
167	NI.070516.182942	WG573662-07	Matrix Spike	20/50	1	L16060907-15	07/05/16 18:29
168	NI.070516.183254	WG573662-06	Matrix Spike Duplica	20/50	1	L16060907-14	07/05/16 18:32
169	NI.070516.183607	WG575179-44	CCV		1		07/05/16 18:36
170	NI.070516.183919	WG575179-45	CCB		1		07/05/16 18:39

Page: 5 Approved: July 06, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070516A.REP

Analyst1: JYH Analyst2: N/A

Method: 6020/6020A/200.8 SOP: ME700A Rev: 2

Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD76567

ICSA: STD76453 ICSAB: STD76573 Int. Std: RGT36607

CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989

Stannous : Hydroxylamine :

Workgroups: 575049,575057,575052,575054,574236,574690,574572,575217

Comments: Additional Workgroup: 574822

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
171	NI.070516.184232	WG573662-08	Matrix Spike Duplica	20/50	1	L16060907-15	07/05/16 18:42
172	NI.070516.184543	L16060907-21	MW36-061516	20/50	1		07/05/16 18:45
173	NI.070516.184855	L16060907-22	MW36-061516	20/50	1		07/05/16 18:48
174	NI.070516.185206	L16061128-01	MW-23-104	20/50	1		07/05/16 18:52
175	NI.070516.185518	L16061128-02	MW-23-103	20/50	1		07/05/16 18:55
176	NI.070516.185829	L16061128-03	FD062016	20/50	1		07/05/16 18:58
177	NI.070516.190141	L16061128-04	EB WATER A	20/50	1		07/05/16 19:01
178	NI.070516.190452	L16061128-04	EB WATER A		100		07/05/16 19:04
179	NI.070516.190804	WG575179-46	Interference Check		1		07/05/16 19:08
180	NI.070516.191116	WG575179-47	Interference Check		1		07/05/16 19:11
181	NI.070516.191429	WG575179-48	CCV		1		07/05/16 19:14
182	NI.070516.191740	WG575179-49	CCB		1		07/05/16 19:17
183	NI.070516.192053	WG575179-50	Low Level Continuing Calibra		1		07/05/16 19:20

Comments

Seq.	Rerun	Dil.	Reason	Analytes
7			Rerun to verify. JYH	
45			Wrong dilution factor. JYH	
90			Wrong sample label. JYH	
100			Seq. 100-102: wrong WG numbers. JYH	
159			Wrong sample label. JYH	

Page: 6 Approved: July 06, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070616B.REP
Analyst1: JYH Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD73705
ICSA: STD77006 ICSAB: STD76573 Int. Std: RGT36607
CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989
Stannous : Hydroxylamine :

Workgroups: 574825,575052,575023,574822,574572,574573,574824,574688

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
1	NI.070616.090452	Blank	Blank		1		07/06/16 09:04
2	NI.070616.090803	WG575370-01	Calibration Point		1		07/06/16 09:08
3	NI.070616.091115	WG575370-02	Calibration Point		1		07/06/16 09:11
4	NI.070616.091426	WG575370-03	Calibration Point		1		07/06/16 09:14
5	NI.070616.091738	WG575370-04	Calibration Point		1		07/06/16 09:17
6	NI.070616.092051	WG575370-05	Initial Calibration Verification		1		07/06/16 09:20
7	NI.070616.092404	WG575370-06	Initial Calib Blank		1		07/06/16 09:24
8	NI.070616.092717	WG575370-07	Low Level Initial Calibration V		1		07/06/16 09:27
9	NI.070616.093029	WG575370-08	Interference Check		1		07/06/16 09:30
10	NI.070616.093340	WG575370-09	Interference Check		1		07/06/16 09:33
11	NI.070616.093653	WG575370-10	CCV		1		07/06/16 09:36
12	NI.070616.094005	WG575370-11	CCB		1		07/06/16 09:40
13	NI.070616.094317	WG574195-02	Method/Prep Blank	.25/100	1		07/06/16 09:43
14	NI.070616.094629	WG574195-03	Laboratory Control S	.25/100	1		07/06/16 09:46
15	NI.070616.094940	WG574195-01	Reference Sample		1	L16061294-11	07/06/16 09:49
16	NI.070616.095252	WG574195-04	Matrix Spike	.257/100	1	L16061294-11	07/06/16 09:52
17	NI.070616.095603	WG574195-05	Matrix Spike Duplica	.252/100	1	L16061294-11	07/06/16 09:56
18	NI.070616.095915	L16061294-01	XU936DS00700A	.255/100	1		07/06/16 09:59
19	NI.070616.100226	L16061294-02	XU936DS00705A	.251/100	1		07/06/16 10:02
20	NI.070616.100538	WG574825-01	Post Digestion Spike		1	L16061294-02	07/06/16 10:05
21	NI.070616.100849	WG574825-02	Serial Dilution		5	L16061294-02	07/06/16 10:08
22	NI.070616.101201	WG574825-02	Serial Dilution		25	L16061294-02	07/06/16 10:12
23	NI.070616.101514	WG575370-12	CCV		1		07/06/16 10:15
24	NI.070616.101825	WG575370-13	CCB		1		07/06/16 10:18
25	NI.070616.102139	L16061294-04	XU936DS01100A	.25/100	1		07/06/16 10:21
26	NI.070616.102450	L16061294-05	XU936DS01105A	.252/100	1		07/06/16 10:24
27	NI.070616.102802	L16061294-06	XU936DS01500A	.25/100	1		07/06/16 10:28
28	NI.070616.103113	L16061294-07	XU936DS01505A	.25/100	1		07/06/16 10:31
29	NI.070616.103425	L16061294-08	XU936DS01505C	.255/100	1		07/06/16 10:34
30	NI.070616.103737	L16061294-09	XU936DS01900A	.256/100	1		07/06/16 10:37
31	NI.070616.104048	L16061294-10	XU936DS01905A	.251/100	1		07/06/16 10:40
32	NI.070616.104400	L16061294-14	XU936DS00200A	.256/100	1		07/06/16 10:44
33	NI.070616.104711	L16061294-15	XU936DS00205A	.251/100	1		07/06/16 10:47
34	NI.070616.105023	L16061381-08	XU936DS00405A	.256/100	1		07/06/16 10:50

Page: 1 Approved: July 07, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070616B.REP
Analyst1: JYH Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD73705
ICSA: STD77006 ICSAB: STD76573 Int. Std: RGT36607
CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989
Stannous : Hydroxylamine :

Workgroups: 574825,575052,575023,574822,574572,574573,574824,574688

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
35	NI.070616.105336	WG575370-14	CCV		1		07/06/16 10:53
36	NI.070616.114231	WG575370-15	CCB		1		07/06/16 11:42
37	NI.070616.114547	WG575370-16	CCV		1		07/06/16 11:45
38	NI.070616.114858	WG575370-17	CCB		1		07/06/16 11:48
39	NI.070616.115343	WG574195-01	Reference Sample		50	L16061294-11	07/06/16 11:53
40	NI.070616.115654	WG574195-04	Matrix Spike	.257/100	50	L16061294-11	07/06/16 11:56
41	NI.070616.120005	WG574195-05	Matrix Spike Duplica	.252/100	50	L16061294-11	07/06/16 12:00
42	NI.070616.120317	L16061294-01	XU936DS00700A	.255/100	50		07/06/16 12:03
43	NI.070616.120628	L16061294-02	XU936DS00705A	.251/100	50		07/06/16 12:06
44	NI.070616.120940	WG574825-01	Post Digestion Spike		50	L16061294-02	07/06/16 12:09
45	NI.070616.121251	WG574825-02	Serial Dilution		250	L16061294-02	07/06/16 12:12
46	NI.070616.121603	L16061294-04	XU936DS01100A	.25/100	50		07/06/16 12:16
47	NI.070616.121914	L16061294-05	XU936DS01105A	.252/100	50		07/06/16 12:19
48	NI.070616.122225	L16061294-06	XU936DS01500A	.25/100	50		07/06/16 12:22
49	NI.070616.122538	WG575370-18	CCV		1		07/06/16 12:25
50	NI.070616.122850	WG575370-19	CCB		1		07/06/16 12:28
51	NI.070616.123202	L16061294-07	XU936DS01505A	.25/100	50		07/06/16 12:32
52	NI.070616.123514	L16061294-08	XU936DS01505C	.255/100	50		07/06/16 12:35
53	NI.070616.123825	L16061294-09	XU936DS01900A	.256/100	50		07/06/16 12:38
54	NI.070616.124137	L16061294-10	XU936DS01905A	.251/100	50		07/06/16 12:41
55	NI.070616.124448	L16061294-14	XU936DS00200A	.256/100	50		07/06/16 12:44
56	NI.070616.124759	L16061294-15	XU936DS00205A	.251/100	50		07/06/16 12:47
57	NI.070616.125111	L16061381-08	XU936DS00405A	.256/100	50		07/06/16 12:51
58	NI.070616.125423	WG575370-20	CCV		1		07/06/16 12:54
59	NI.070616.125735	WG575370-21	CCB		1		07/06/16 12:57
60	NI.070616.130440	L16061313-06	MW-4	20/50	1		07/06/16 13:04
61	NI.070616.130752	L16061313-07	MW-4	20/50	1		07/06/16 13:07
62	NI.070616.131103	L16061313-08	MW-3	20/50	1		07/06/16 13:11
63	NI.070616.131415	L16061313-09	MW-3	20/50	1		07/06/16 13:14
64	NI.070616.131726	L16061313-10	DUPLICATE	20/50	1		07/06/16 13:17
65	NI.070616.132037	L16061313-11	DUPLICATE	20/50	1		07/06/16 13:20
66	NI.070616.132349	L16061313-12	MW-2D	20/50	1		07/06/16 13:23
67	NI.070616.132700	L16061313-13	MW-2D	20/50	1		07/06/16 13:27
68	NI.070616.133012	L16061313-14	EQ-RINSE	20/50	1		07/06/16 13:30

Page: 2 Approved: July 07, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070616B.REP
Analyst1: JYH Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD73705
ICSA: STD77006 ICSAB: STD76573 Int. Std: RGT36607
CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989
Stannous : Hydroxylamine :

Workgroups: 574825,575052,575023,574822,574572,574573,574824,574688

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
69	NI.070616.133325	WG575370-22	CCV		1		07/06/16 13:33
70	NI.070616.133636	WG575370-23	CCB		1		07/06/16 13:36
71	NI.070616.134017	WG574881-02	Method/Prep Blank	.25/100	1		07/06/16 13:40
72	NI.070616.134329	WG574881-03	Laboratory Control S	.25/100	1		07/06/16 13:43
73	NI.070616.134640	WG574881-01	Reference Sample		1	L16061533-03	07/06/16 13:46
74	NI.070616.134951	WG574881-04	Matrix Spike	.25/100	1	L16061533-03	07/06/16 13:49
75	NI.070616.135303	WG574881-05	Matrix Spike Duplica	.254/100	1	L16061533-03	07/06/16 13:53
76	NI.070616.135614	L16061488-01	NV79-01	.25/100	1		07/06/16 13:56
77	NI.070616.135926	L16061488-02	NV79-02	.258/100	1		07/06/16 13:59
78	NI.070616.140318	WG575023-01	Post Digestion Spike		1	L16061488-02	07/06/16 14:03
79	NI.070616.140630	WG575023-02	Serial Dilution		5	L16061488-02	07/06/16 14:06
80	NI.070616.140941	WG575023-02	Serial Dilution		25	L16061488-02	07/06/16 14:09
81	NI.070616.141254	WG575370-24	CCV		1		07/06/16 14:12
82	NI.070616.141606	WG575370-25	CCB		1		07/06/16 14:16
83	NI.070616.142024	L16061488-03	NV79-03	.259/100	1		07/06/16 14:20
84	NI.070616.142336	L16061488-04	NV79-04	.257/100	1		07/06/16 14:23
85	NI.070616.142647	L16061533-01	TS935DS01100A	.253/100	1		07/06/16 14:26
86	NI.070616.142958	L16061533-02	TS935DS01100C	.255/100	1		07/06/16 14:29
87	NI.070616.143310	L16061643-01	V6F0315-01	.254/100	1		07/06/16 14:33
88	NI.070616.143621	L16061643-02	V6F0315-02	.253/100	1		07/06/16 14:36
89	NI.070616.143933	L16061687-01	TR0059	.252/100	1		07/06/16 14:39
90	NI.070616.144244	L16061296-01	OUTFALL 002/COMP		1		07/06/16 14:42
91	NI.070616.144556	L16061313-08	MW-3	20/50	10		07/06/16 14:45
92	NI.070616.144907	L16061313-10	DUPLICATE	20/50	10		07/06/16 14:49
93	NI.070616.145220	WG575370-26	CCV		1		07/06/16 14:52
94	NI.070616.145531	WG575370-27	CCB		1		07/06/16 14:55
95	NI.070616.150731	L16060907-08	MW11S-061516	20/50	100		07/06/16 15:07
96	NI.070616.151042	L16060907-10	DUP-GW-061516	20/50	100		07/06/16 15:10
97	NI.070616.151354	L16060907-11	DUP-GW-061516	20/50	100		07/06/16 15:13
98	NI.070616.151705	L16060907-12	MW31-061516	20/50	1		07/06/16 15:17
99	NI.070616.152017	L16060907-13	MW31-061516	20/50	1		07/06/16 15:20
100	NI.070616.152328	L16061026-02	18WW16F-061616		1		07/06/16 15:23
101	NI.070616.152752	L16061026-02	18WW16F-061616	20/50	200		07/06/16 15:27
102	NI.070616.153105	WG575370-28	CCV		1		07/06/16 15:31

Page: 3 Approved: July 07, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070616B.REP
Analyst1: JYH Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD73705
ICSA: STD77006 ICSAB: STD76573 Int. Std: RGT36607
CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989
Stannous : Hydroxylamine :

Workgroups: 574825,575052,575023,574822,574572,574573,574824,574688

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
103	NI.070616.153417	WG575370-29	CCB		1		07/06/16 15:34
104	NI.070616.155033	WG574306-02	Method/Prep Blank	.25/100	1		07/06/16 15:50
105	NI.070616.155345	WG574306-03	Laboratory Control S	.25/100	1		07/06/16 15:53
106	NI.070616.155656	L16061297-01	XU936DS01805A	.259/100	1		07/06/16 15:56
107	NI.070616.160008	L16061297-02	XU936DS01800A	.25/100	1		07/06/16 16:00
108	NI.070616.160319	L16061297-03	XU936DS01405A	.256/100	1		07/06/16 16:03
109	NI.070616.160631	L16061297-04	XU936DS01405C	.258/100	1		07/06/16 16:06
110	NI.070616.160942	L16061297-05	XU936DS02005A	.25/100	1		07/06/16 16:09
111	NI.070616.161318	L16061297-06	XU936DS01605A	.258/100	1		07/06/16 16:13
112	NI.070616.161629	WG574573-03	Post Digestion Spike		1	L16061297-06	07/06/16 16:16
113	NI.070616.161940	WG574573-04	Serial Dilution		5	L16061297-06	07/06/16 16:19
114	NI.070616.162253	WG575370-30	CCV		1		07/06/16 16:22
115	NI.070616.162605	WG575370-31	CCB		1		07/06/16 16:26
116	NI.070616.162917	WG574306-01	Reference Sample		1	L16061297-07	07/06/16 16:29
117	NI.070616.163229	WG574306-04	Matrix Spike	.253/100	1	L16061297-07	07/06/16 16:32
118	NI.070616.163540	WG574306-05	Matrix Spike Duplica	.255/100	1	L16061297-07	07/06/16 16:35
119	NI.070616.163852	L16061297-10	XU936DS01705A	.254/100	1		07/06/16 16:38
120	NI.070616.164203	L16061297-11	XU936DS01305A	.253/100	1		07/06/16 16:42
121	NI.070616.164515	L16061297-12	XU936DS01205A	.251/100	1		07/06/16 16:45
122	NI.070616.164826	L16061297-13	XU936DS00800A	.252/100	1		07/06/16 16:48
123	NI.070616.165137	L16061297-14	XU936DS00805A	.25/100	1		07/06/16 16:51
124	NI.070616.165449	L16061297-15	XU936DS00900A	.259/100	1		07/06/16 16:54
125	NI.070616.165800	L16061297-16	XU936DS00905A	.256/100	1		07/06/16 16:58
126	NI.070616.170113	WG575370-32	CCV		1		07/06/16 17:01
127	NI.070616.170425	WG575370-33	CCB		1		07/06/16 17:04
128	NI.070616.170737	L16061297-17	XU936DS00905C	.25/100	1		07/06/16 17:07
129	NI.070616.171049	L16061297-18	XU936DS01005A	.255/100	1		07/06/16 17:10
130	NI.070616.171400	L16061297-19	XU936DS00600A	.258/100	1		07/06/16 17:14
131	NI.070616.171712	L16061297-20	XU936DS00605A	.259/100	1		07/06/16 17:17
132	NI.070616.172023	L16061297-01	XU936DS01805A	.259/100	50		07/06/16 17:20
133	NI.070616.172334	L16061297-02	XU936DS01800A	.25/100	50		07/06/16 17:23
134	NI.070616.172646	L16061297-03	XU936DS01405A	.256/100	50		07/06/16 17:26
135	NI.070616.172957	L16061297-04	XU936DS01405C	.258/100	50		07/06/16 17:29
136	NI.070616.173309	L16061297-05	XU936DS02005A	.25/100	50		07/06/16 17:33

Page: 4 Approved: July 07, 2016

Microbac Laboratories Inc.

Instrument Run Log

Instrument: ICP-MS2 Dataset: 070616B.REP
Analyst1: JYH Analyst2: N/A
Method: 6020/6020A/200.8 SOP: ME700A Rev: 2
Maintenance Log ID:

Calibration Std: STD76710 ICV Std: STD76906 Post Spike: STD73705
ICSA: STD77006 ICSAB: STD76573 Int. Std: RGT36607
CCV: STD76905 LLCCV: STD76908 Tuning Sol : STD76989
Stannous : Hydroxylamine :

Workgroups: 574825,575052,575023,574822,574572,574573,574824,574688

Comments:

Seq.	File ID	Sample	ID	Prep	Dil	Reference	Date/Time
137	NI.070616.173620	L16061297-06	XU936DS01605A	.258/100	50		07/06/16 17:36
138	NI.070616.173934	WG575370-34	CCV		1		07/06/16 17:39
139	NI.070616.174245	WG575370-35	CCB		1		07/06/16 17:42
140	NI.070616.174559	WG574573-03	Post Digestion Spike		50	L16061297-06	07/06/16 17:45
141	NI.070616.174910	WG574573-04	Serial Dilution		250	L16061297-06	07/06/16 17:49
142	NI.070616.175222	WG574306-01	Reference Sample		50	L16061297-07	07/06/16 17:52
143	NI.070616.175533	WG574306-04	Matrix Spike	.253/100	50	L16061297-07	07/06/16 17:55
144	NI.070616.175844	WG574306-05	Matrix Spike Duplica	.255/100	50	L16061297-07	07/06/16 17:58
145	NI.070616.180156	L16061297-10	XU936DS01705A	.254/100	50		07/06/16 18:01
146	NI.070616.180507	L16061297-11	XU936DS01305A	.253/100	50		07/06/16 18:05
147	NI.070616.180818	L16061297-12	XU936DS01205A	.251/100	50		07/06/16 18:08
148	NI.070616.181130	L16061297-13	XU936DS00800A	.252/100	50		07/06/16 18:11
149	NI.070616.181441	L16061297-14	XU936DS00805A	.25/100	50		07/06/16 18:14
150	NI.070616.181755	WG575370-36	CCV		1		07/06/16 18:17
151	NI.070616.182106	WG575370-37	CCB		1		07/06/16 18:21
152	NI.070616.182419	L16061297-15	XU936DS00900A	.259/100	50		07/06/16 18:24
153	NI.070616.182731	L16061297-16	XU936DS00905A	.256/100	50		07/06/16 18:27
154	NI.070616.183042	L16061297-17	XU936DS00905C	.25/100	50		07/06/16 18:30
155	NI.070616.183353	L16061297-18	XU936DS01005A	.255/100	50		07/06/16 18:33
156	NI.070616.183705	L16061297-19	XU936DS00600A	.258/100	50		07/06/16 18:37
157	NI.070616.184016	L16061297-20	XU936DS00605A	.259/100	50		07/06/16 18:40
158	NI.070616.184328	L16061452-01	18CPTMW12SW-062416	20/50	50		07/06/16 18:43
159	NI.070616.184639	L16061453-01	MW2-062416	20/50	50		07/06/16 18:46
160	NI.070616.184952	WG575370-38	CCV		1		07/06/16 18:49
161	NI.070616.185304	WG575370-39	CCB		1		07/06/16 18:53
162	NI.070616.185616	L16061453-03	18CPTMW01SW-062416	20/50	50		07/06/16 18:56
163	NI.070616.185928	L16061296-01	OUTFALL 002/COMP	20/50	1		07/06/16 18:59
164	NI.070616.190239	L16061453-03	18CPTMW01SW-062416		500		07/06/16 19:02
165	NI.070616.190551	WG575370-40	Interference Check		1		07/06/16 19:05
166	NI.070616.190902	WG575370-41	Interference Check		1		07/06/16 19:09
167	NI.070616.191215	WG575370-42	CCV		1		07/06/16 19:12
168	NI.070616.191526	WG575370-43	CCB		1		07/06/16 19:15
169	NI.070616.191839	WG575370-44	Low Level Continuing Calibra		1		07/06/16 19:18

Page: 5 Approved: July 07, 2016

Microbac Laboratories Inc.

Data Checklist

Date: 29-JUN-2016
Analyst: BKT
Analyst: NA
Method: 6020/6020A/200.8
Instrument: ICP-MS2
Curve Workgroup: 574676
Runlog ID: 76013
Analytical Workgroups: 573722,574570,574572,573717,574688,574690,573790

STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	
ICB/CCB	X
ICSA/ICSAB	X
CRI	
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	X
Client Forms	X
Level X	
Level 3	1279,1294,1381,1384
Level 4	661,822,823,883,912,974,986,1026
Level 4	1206,1207,1208,1265,1266,1269,1272
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	BKT
Secondary Reviewer	JYH
Comments	

Primary Reviewer:
04-JUL-2016

Secondary Reviewer:
05-JUL-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUL-05-2016 12:36:33

Microbac Laboratories Inc.

Data Checklist

Date: 05-JUL-2016
Analyst: JYH
Analyst: NA
Method: 6020/6020A/200.8
Instrument: ICP-MS2
Curve Workgroup: 575179
Runlog ID: 76106
Analytical Workgroups: 575049,575057,575052,575054,574236,574690,574572,575217

Additional Workgroup	574822
STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	
ICB/CCB	X
ICSA/ICSAB	X
CRI	
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	X
Client Forms	X
Level X	
Level 3	1445,1533,1128
Level 4	1451,1559,840,1026,088,907
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	JYH
Secondary Reviewer	BKT
Comments	

Primary Reviewer:
06-JUL-2016

Secondary Reviewer:
06-JUL-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUL-06-2016 20:27:37

Microbac Laboratories Inc.

Data Checklist

Date: 06-JUL-2016
Analyst: JYH
Analyst: NA
Method: 6020/6020A/200.8
Instrument: ICP-MS2
Curve Workgroup: 575370
Runlog ID: 76130
Analytical Workgroups: 574825,575052,575023,574822,574572,574573,574824,574688

STD ID#s on Runlog	X
Calibration/Linearity	X
ICV/CCV	X
ICV RSD < 3% (EPA 200.7 only)	
ICB/CCB	X
ICSA/ICSAB	X
CRI	
Blank/LCS	X
MS/MSD	X
Post Spike/Serial Dilution	X
Upload Results	X
Data Qualifiers	
Generate PDF Instrument Data	X
Sign/Annotate PDF Data	X
Upload Curve Data	X
Workgroup Forms	X
Case Narrative	X
Client Forms	X
Level X	
Level 3	1294,1533,1297
Level 4	1687,907,1026,1452,1453
Check for compliance with method and project specific requirements	X
Check the completeness of reported information	X
Check the information for the report narrative	X
Primary Reviewer	JYH
Secondary Reviewer	BKT
Comments	

Primary Reviewer:
07-JUL-2016

Secondary Reviewer:
07-JUL-2016

CHECKLIST1 - Modified 03/05/2008
Generated: JUL-07-2016 15:29:19

HOLDING TIMES
EQUIVALENT TO AFCEE FORM 9

Analytical Method: 6020A

AAB#: WG574572

Login Number: L16061026

Client ID	ID	Date Collected	TCLP Date	Time Held	Max Hold	Q	Extract Date	Time Held	Max Hold	Q	Run Date	Time Held	Max Hold	Q
18WW16F-061616	02	06/16/16					06/21/2016	5.1	180		07/06/16	20.3	180	
18WW16F-061616	02	06/16/16					06/21/2016	5.1	180		07/05/16	19.3	180	
18WW16F-061616	02	06/16/16					06/21/2016	5.1	180		06/29/16	13.3	180	
18WW16F-061616	02	06/16/16					06/21/2016	5.1	180		07/05/16	19.3	180	
129-061616	04	06/16/16					06/21/2016	5.1	180		07/05/16	19.3	180	
129-061616	04	06/16/16					06/21/2016	5.1	180		07/05/16	19.3	180	
129-061616	04	06/16/16					06/21/2016	5.1	180		06/29/16	13.2	180	
120F-061616	06	06/16/16					06/21/2016	5.1	180		06/29/16	13.2	180	
120F-061616	06	06/16/16					06/21/2016	5.1	180		07/05/16	19.3	180	
123-061616	07	06/16/16					06/21/2016	5.1	180		06/29/16	13.2	180	
123-061616	07	06/16/16					06/21/2016	5.1	180		07/05/16	19.3	180	

* = SEE PROJECT QAPP REQUIREMENTS

HOLD_TIMES - Modified 03/06/2008
PDF File ID: 4832316
Report generated 07/07/2016 08:33

METHOD BLANK SUMMARY

Login Number: L16061026
Blank File ID: NI.062916.125354
Prep Date: 06/21/16 11:18
Analyzed Date: 06/29/16 12:53
Analyst: BKT

Work Group: WG574572
Blank Sample ID: WG573548-02
Instrument ID: ICP-MS2
Method: 6020A

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573548-03	NI.062916.125705	06/29/16 12:57	01
18WW16F-061616	L16061026-02	NI.062916.145120	06/29/16 14:51	01
129-061616	L16061026-04	NI.062916.145431	06/29/16 14:54	01
120F-061616	L16061026-06	NI.062916.145742	06/29/16 14:57	01
123-061616	L16061026-07	NI.062916.150054	06/29/16 15:00	01
LCS	WG573548-03	NI.070516.155247	07/05/16 15:52	02
18WW16F-061616	L16061026-02	NI.070516.160533	07/05/16 16:05	DL01
18WW16F-061616	L16061026-02	NI.070516.160845	07/05/16 16:08	DL02
129-061616	L16061026-04	NI.070516.161156	07/05/16 16:11	02
129-061616	L16061026-04	NI.070516.161508	07/05/16 16:15	DL01
120F-061616	L16061026-06	NI.070516.163034	07/05/16 16:30	02
123-061616	L16061026-07	NI.070516.163440	07/05/16 16:34	02
18WW16F-061616	L16061026-02	NI.070616.152752	07/06/16 15:27	DL03

Report Name: BLANK_SUMMARY
PDF File ID: 4832317
Report generated 07/07/2016 08:33

METHOD BLANK SUMMARY

Login Number: L16061026
Blank File ID: NI.070516.154936
Prep Date: 06/21/16 11:18
Analyzed Date: 07/05/16 15:49
Analyst: JYH

Work Group: WG574572
Blank Sample ID: WG573548-02
Instrument ID: ICP-MS2
Method: 6020A

This Method Blank Applies To The Following Samples:

Client ID	Lab Sample ID	Lab File ID	Time Analyzed	TAG
LCS	WG573548-03	NI.062916.125705	06/29/16 12:57	01
18WW16F-061616	L16061026-02	NI.062916.145120	06/29/16 14:51	01
129-061616	L16061026-04	NI.062916.145431	06/29/16 14:54	01
120F-061616	L16061026-06	NI.062916.145742	06/29/16 14:57	01
123-061616	L16061026-07	NI.062916.150054	06/29/16 15:00	01
LCS	WG573548-03	NI.070516.155247	07/05/16 15:52	02
18WW16F-061616	L16061026-02	NI.070516.160533	07/05/16 16:05	DL01
18WW16F-061616	L16061026-02	NI.070516.160845	07/05/16 16:08	DL02
129-061616	L16061026-04	NI.070516.161156	07/05/16 16:11	02
129-061616	L16061026-04	NI.070516.161508	07/05/16 16:15	DL01
120F-061616	L16061026-06	NI.070516.163034	07/05/16 16:30	02
123-061616	L16061026-07	NI.070516.163440	07/05/16 16:34	02
18WW16F-061616	L16061026-02	NI.070616.152752	07/06/16 15:27	DL03

Report Name: BLANK_SUMMARY
PDF File ID: 4832317
Report generated 07/07/2016 08:33

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16061026 Prep Date: 06/21/16 11:18 Sample ID: WG573548-02
Instrument ID: ICP-MS2 Run Date: 06/29/16 12:53 Prep Method: 3015
File ID: NI.062916.125354 Analyst: BKT Method: 6020A
Workgroup (AAB#): WG574572 Matrix: Water Units: mg/L
Contract #: Cal ID: ICP-MS-29-JUN-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Antimony, Total	0.000500	0.00200	0.000500	1	U
Arsenic, Total	0.000500	0.00200	0.000500	1	U
Barium, Total	0.00150	0.00600	0.00150	1	U
Cadmium, Total	0.000300	0.00120	0.000300	1	U
Chromium, Total	0.00100	0.00400	0.00100	1	U
Copper, Total	0.00100	0.00400	0.00100	1	U
Lead, Total	0.000500	0.00200	0.000500	1	U
Manganese, Total	0.00100	0.00400	0.00239	1	*
Nickel, Total	0.00200	0.00800	0.00200	1	U
Silver, Total	0.000500	0.00200	0.000589	1	J
Thallium, Total	0.000100	0.000400	0.000100	1	U
Vanadium, Total	0.000500	0.00200	0.000500	1	U
Zinc, Total	0.0125	0.0500	0.0125	1	U

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4832318
07-JUL-2016 08:33

Microbac Laboratories Inc.
METHOD BLANK REPORT

Login Number: L16061026 Prep Date: 06/21/16 11:18 Sample ID: WG573548-02
Instrument ID: ICP-MS2 Run Date: 07/05/16 15:49 Prep Method: 3015
File ID: NI.070516.154936 Analyst: JYH Method: 6020A
Workgroup (AAB#): WG574572 Matrix: Water Units: mg/L
Contract #: Cal ID: ICP-MS-05-JUL-16

Analytes	DL	LOQ	Concentration	Dilution	Qualifier
Cobalt, Total	0.000500	0.00200	0.000500	1	U

DL Method Detection Limit
LOQ Reporting/Practical Quantitation Limit
ND Analyte Not detected at or above reporting limit
* |Analyte concentration| > 1/2 RL

Report Name: BLANK
PDF ID: 4832318
07-JUL-2016 08:33

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16061026 Run Date: 06/29/2016 Sample ID: WG573548-03
Instrument ID: ICP-MS2 Run Time: 12:57 Prep Method: 3015
File ID: NI.062916.125705 Analyst: BKT Method: 6020A
Workgroup (AAB#): WG574572 Matrix: Water Units: mg/L
QC Key: DOD4 Lot#: STD75998 Cal ID: ICP-MS-29-JUN-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Antimony, Total	0.125	0.118	94.4	80 - 120	
Arsenic, Total	0.125	0.122	97.5	80 - 120	
Barium, Total	0.125	0.121	97.0	80 - 120	
Cadmium, Total	0.125	0.121	96.8	80 - 120	
Chromium, Total	0.125	0.130	104	80 - 120	
Copper, Total	0.125	0.131	105	80 - 120	
Lead, Total	0.125	0.123	98.3	80 - 120	
Manganese, Total	0.125	0.136	109	80 - 120	
Nickel, Total	0.125	0.129	104	80 - 120	
Silver, Total	0.125	0.123	98.7	80 - 120	
Thallium, Total	0.125	0.123	98.1	80 - 120	
Vanadium, Total	0.125	0.127	102	80 - 120	
Zinc, Total	0.125	0.134	107	80 - 120	

LCS - Modified 03/06/2008
PDF File ID: 4832319
Report generated: 07/07/2016 08:33

Microbac Laboratories Inc.
LABORATORY CONTROL SAMPLE (LCS)

Login Number: L16061026 Run Date: 07/05/2016 Sample ID: WG573548-03
Instrument ID: ICP-MS2 Run Time: 15:52 Prep Method: 3015
File ID: NI.070516.155247 Analyst: JYH Method: 6020A
Workgroup (AAB#): WG574572 Matrix: Water Units: mg/L
QC Key: DOD4 Lot#: STD75998 Cal ID: ICP-MS-05-JUL-16

Analytes	Expected	Found	% Rec	LCS Limits	Q
Cobalt, Total	0.125	0.115	91.6	80 - 120	

LCS - Modified 03/06/2008
PDF File ID: 4832319
Report generated: 07/07/2016 08:33

Microbac Laboratories Inc.
MATRIX SPIKE AND MATRIX SPIKE DUP (MS/MSD)

Loginnum: L16061026 Cal ID: ICP-MS2- Worknum: WG574572
Instrument ID: ICP-MS2 Contract #: Method: 6020A
Parent ID: WG573548-01 File ID: NI.062916.130017 Dil: 1 Matrix: WATER
Sample ID: WG573548-04 MS File ID: NI.062916.130328 Dil: 1 Units: mg/L
Sample ID: WG573548-05 MSD File ID: NI.062916.130639 Dil: 1

Analyte	Parent	MS Spiked	MS Found	MS %Rec	MSD Spiked	MSD Found	MSD %Rec	%RPD	%Rec Limits	RPD Limit	Q
Antimony	0.000929	0.125	0.123	97.5	0.125	0.126	100	2.42	80 - 120	20	
Arsenic	0.00184	0.125	0.137	108	0.125	0.136	108	0.245	80 - 120	20	
Barium	0.0277	0.125	0.147	95.5	0.125	0.149	96.9	1.17	80 - 120	20	
Cadmium	ND	0.125	0.121	96.7	0.125	0.123	98.3	1.66	80 - 120	20	
Chromium	0.00161	0.125	0.127	100	0.125	0.125	98.5	1.86	80 - 120	20	
Copper	0.0278	0.125	0.160	106	0.125	0.135	85.4	17.4	80 - 120	20	
Lead	ND	0.125	0.125	99.9	0.125	0.129	103	3.34	80 - 120	20	
Manganese	0.572	0.125	0.687	92.1	0.125	0.694	97.6	0.980	80 - 120	20	
Nickel	0.00762	0.125	0.130	97.9	0.125	0.130	97.8	0.0660	80 - 120	20	
Silver	ND	0.125	0.0534	42.7	0.125	0.0494	39.5	7.77	80 - 120	20	*
Thallium	ND	0.125	0.125	100	0.125	0.126	101	0.495	80 - 120	20	
Vanadium	ND	0.125	0.133	106	0.125	0.134	107	0.902	80 - 120	20	
Zinc	0.0139	0.125	0.143	104	0.125	0.133	95.7	7.23	80 - 120	20	

* FAILS %REC LIMIT

FAILS RPD LIMIT

NOTE: This is an internal quality control sample.

WG_MS_MSD_DRYWT - Modified 05/26/2011
PDF File ID: 4832320
Report generated 07/07/2016 08:33

Microbac Laboratories Inc.
MATRIX SPIKE AND MATRIX SPIKE DUP (MS/MSD)

Loginnum: L16061026 Cal ID: ICP-MS2 - Worknum: WG574572
Instrument ID: ICP-MS2 Contract #: Method: 6020A
Parent ID: WG573548-01 File ID: NI.070516.155559 Dil: 1 Matrix: WATER
Sample ID: WG573548-04 MS File ID: NI.070516.155910 Dil: 1 Units: mg/L
Sample ID: WG573548-05 MSD File ID: NI.070516.160221 Dil: 1

Analyte	Parent	MS Spiked	MS Found	MS %Rec	MSD Spiked	MSD Found	MSD %Rec	%RPD	%Rec Limits	RPD Limit	Q
Cobalt	0.000653	0.125	0.122	97.1	0.125	0.118	94.0	3.22	80 - 120	20	

* FAILS %REC LIMIT

FAILS RPD LIMIT

NOTE: This is an internal quality control sample.

WG_MS_MSD_DRYWT - Modified 05/26/2011
PDF File ID: 4832320
Report generated 07/07/2016 08:33

Microbac Laboratories Inc.
Serial Dilution Report

Login: L16061026 **Worknum:** WG574572
Instrument: ICP-MS2 **Method:** 6020A
Serial Dil: WG574572-02 **File ID:** NI.062916.131925 **Dil:** 5 **Units:** ug/L
Sample: L16060986-03 **File ID:** NI.062916.131302 **Dil:** 1

Analyte	Sample	Qual	Serial Dil	Qual	% Diff	Q
Antimony	ND	U	2.02	F	1290.00	
Arsenic	1.45	X	1.60	F	10.50	
Barium	11.5	X	11.5	F	0.05	
Cadmium	ND	U	ND	U		
Chromium	0.585	F	ND	U		
Cobalt	2.82	X	2.79	F	1.29	
Copper	8.96	X	9.10	X	1.62	
Lead	ND	U	ND	U		
Manganese	1780		1750		1.52	
Nickel	6.40	X	6.54	F	2.06	
Silver	ND	U	ND	U		
Thallium	ND	U	ND	U		
Vanadium	0.203	F	ND	U		
Zinc	22.4	X	27.8	F	24.50	

U = Result is below MDL.
F = Result is greater than or equal to MDL and less than the RL.
X = Result is greater than or equal to RL and less than 100 times the MDL.
E = %D exceeds control limit of 10% and initial sample result is greater than or equal to 100 times the MDL.

SERIAL_DIL - Modified 09/22/2008
PDF File ID: 4832314
07/07/2016 08:33

Microbac Laboratories Inc.
Serial Dilution Report

Login: L16061026 **Worknum:** WG574572
Instrument: ICP-MS2 **Method:** 6020A
Serial Dil: WG574572-04 **File ID:** NI.070516.164103 **Dil:** 5 **Units:** ug/L
Sample: L16061026-07 **File ID:** NI.070516.163440 **Dil:** 1

Analyte	Sample	Qual	Serial Dil	Qual	% Diff	Q
Antimony	ND	U	1.70	F	1140.00	
Arsenic	0.283	F	ND	U		
Barium	49.6	X	54.2	X	9.24	
Cadmium	ND	U	ND	U		
Chromium	0.588	F	ND	U		
Cobalt	1.00	X	1.09	F	9.13	
Copper	0.617	F	ND	U		
Lead	0.406	F	ND	U		
Manganese	5.52	X	5.72	F	3.64	
Nickel	0.977	F	ND	U		
Silver	ND	U	ND	U		
Thallium	ND	U	ND	U		
Vanadium	0.831	X	ND	U		
Zinc	ND	U	ND	U		

U = Result is below MDL.
F = Result is greater than or equal to MDL and less than the RL.
X = Result is greater than or equal to RL and less than 100 times the MDL.
E = %D exceeds control limit of 10% and initial sample result is greater than or equal to 100 times the MDL.

SERIAL_DIL - Modified 09/22/2008
PDF File ID: 4832314
07/07/2016 08:33

Microbac Laboratories Inc.
POST SPIKE REPORT

Sample Login ID: L16061026 **Worknum:** WG574572
Instrument ID: ICP-MS2 **Method:** 6020A
Post Spike ID: WG574572-01 **File ID:** NI.062916.131614 **Dil:** 1 **Units:** ug/L
Sample ID: L16060986-03 **File ID:** NI.062916.131302 **Dil:** 1 **Matrix:** Water

Analyte	Post Spike Result	C	Sample Result	C	Spike Added(SA)	% R	Control Limit %R	Q
ANTIMONY	49.8		0	U	50	99.5	75 - 125	
ARSENIC	56.4		1.45		50	109.9	75 - 125	
BARIUM	59.4		11.5		50	95.9	75 - 125	
CADMIUM	50.9		0	U	50	101.9	75 - 125	
CHROMIUM	52.3		0.585	F	50	103.5	75 - 125	
COBALT	61.7		2.82		50	117.8	75 - 125	
COPPER	58.9		8.96		50	100.0	75 - 125	
LEAD	52.9		0	U	50	105.9	75 - 125	
MANGANESE	1830		1780		50	91.5	75 - 125	
NICKEL	56.4		6.40		50	100.0	75 - 125	
SILVER	49.7		0	U	50	99.5	75 - 125	
THALLIUM	52.3		0	U	50	104.6	75 - 125	
VANADIUM	56.0		0.203	F	50	111.6	75 - 125	
ZINC	73.9		22.4		50	103.0	75 - 125	

N = % Recovery exceeds control limits
F = Result is between MDL and RL
U = Sample result is below MDL. A value of zero is used in the calculation

POST_SPIKE - Modified 03/06/2008
PDF File ID: 4832315
Report generated: 07/07/2016 08:33

Microbac Laboratories Inc.
POST SPIKE REPORT

Sample Login ID: L16061026 **Worknum:** WG574572
Instrument ID: ICP-MS2 **Method:** 6020A
Post Spike ID: WG574572-03 **File ID:** NI.070516.163751 **Dil:** 1 **Units:** ug/L
Sample ID: L16061026-07 **File ID:** NI.070516.163440 **Dil:** 1 **Matrix:** Water

Analyte	Post Spike Result	C	Sample Result	C	Spike Added(SA)	% R	Control Limit %R	Q
ANTIMONY	47.9		0	U	50	95.8	75 - 125	
ARSENIC	48.7		0.283	F	50	96.8	75 - 125	
BARIUM	95.8		49.6		50	92.3	75 - 125	
CADMIUM	48.2		0	U	50	96.5	75 - 125	
CHROMIUM	47.4		0.588	F	50	93.6	75 - 125	
COBALT	49.1		1.00		50	96.2	75 - 125	
COPPER	48.5		0.617	F	50	95.8	75 - 125	
LEAD	50.1		0.406	F	50	99.3	75 - 125	
MANGANESE	51.9		5.52		50	92.8	75 - 125	
NICKEL	48.0		0.977	F	50	94.0	75 - 125	
SILVER	47.2		0	U	50	94.4	75 - 125	
THALLIUM	47.8		0	U	50	95.6	75 - 125	
VANADIUM	47.9		0.831		50	94.1	75 - 125	
ZINC	52.8		0	U	50	105.7	75 - 125	

N = % Recovery exceeds control limits
F = Result is between MDL and RL
U = Sample result is below MDL. A value of zero is used in the calculation

POST_SPIKE - Modified 03/06/2008
PDF File ID: 4832315
Report generated: 07/07/2016 08:33

Microbac Laboratories Inc.
Initial Calibration Summary

Login:	L16061026	Workgroup (AAB#):	WG574572
Analytical Method:	6020A	Instrument ID:	ICP-MS2
ICAL Worknum:	WG574676	Initial Calibration Date:	29-JUN-2016 11:35

	WG574676-01		WG574676-02		WG574676-03		WG574676-04			
	Conc	INT	Conc	INT	Conc	INT	Conc	INT	R	Q
ANTIMONY	0	28.2	.4	255	50	227000	100	448000	.999981	
ARSENIC	0	-40.0	.4	-10.4	50	57000	100	110000	.999904	
BARIUM	0	25.3	.4	136	50	99600	100	196000	.999974	
CADMIUM	0	1.30	.4	122	50	102000	100	200000	.999946	
CHROMIUM	0	7540	.4	7930	50	379000	100	748000	.999997	
COBALT	0	239	.4	575	50	337000	100	648000	.99988	
COPPER	0	103	.4	285	50	94300	100	184000	.999979	
LEAD	0	873	.4	2280	50	926000	100	1830000	.999949	
MANGANESE	0	784	.4	1380	50	355000	100	686000	.999925	
NICKEL	0	221	.4	324	50	94600	100	186000	.999991	
SILVER	0	86.3	.4	412	50	318000	100	624000	.999961	
THALLIUM	0	1.70	.4	417	50	406000	100	804000	.999963	
VANADIUM	0	1430	.4	1850	50	385000	100	745000	.999935	
ZINC	0	365	.4	549	50	56000	100	108000	.999932	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
*** = Out of Compliance; R < 0.995**

INT_CAL_ICP - Modified 03/06/2008
PDF File ID: 4832323
Report generated: 07-JUL-2016 08:33

Microbac Laboratories Inc.
Initial Calibration Summary

Login:	L16061026	Workgroup (AAB#):	WG574572
Analytical Method:	6020A	Instrument ID:	ICP-MS2
ICAL Worknum:	WG575179	Initial Calibration Date:	05-JUL-2016 09:36

	WG575179-01		WG575179-02		WG575179-03		WG575179-04			
	Conc	INT	Conc	INT	Conc	INT	Conc	INT	R	Q
ANTIMONY	0	109	.4	306	50	221000	100	451000	.999882	
ARSENIC	0	-83.8	.4	-2.10	50	57500	100	112000	.99994	
BARIUM	0	34.0	.4	149	50	99900	100	193000	.999924	
CADMIUM	0	3.00	.4	116	50	106000	100	204000	.999881	
CHROMIUM	0	6920	.4	7550	50	368000	100	707000	.999922	
COBALT	0	156	.4	528	50	345000	100	669000	.999911	
COPPER	0	141	.4	251	50	95800	100	184000	.999865	
LEAD	0	731	.4	1810	50	908000	100	1770000	.999981	
MANGANESE	0	831	.4	1190	50	322000	100	624000	.999917	
NICKEL	0	385	.4	484	50	95600	100	184000	.999886	
SILVER	0	67.3	.4	421	50	337000	100	642000	.999841	
THALLIUM	0	4.30	.4	436	50	404000	100	778000	.999935	
VANADIUM	0	1410	.4	1710	50	367000	100	713000	.99994	
ZINC	0	188	.4	239	50	56400	100	109000	.999877	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
*** = Out of Compliance; R < 0.995**

INT_CAL_ICP - Modified 03/06/2008
PDF File ID: 4832323
Report generated: 07-JUL-2016 08:33

Microbac Laboratories Inc.
Initial Calibration Summary

Login:	L16061026	Workgroup (AAB#):	WG574572
Analytical Method:	6020A	Instrument ID:	ICP-MS2
ICAL Worknum:	WG575370	Initial Calibration Date:	06-JUL-2016 09:17

	WG575370-01		WG575370-02		WG575370-03		WG575370-04			
	Conc	INT	Conc	INT	Conc	INT	Conc	INT	R	Q
ANTIMONY	0	280	.4	308	50	202000	100	429000	.999223	
ARSENIC	0	-74.4	.4	-50.4	50	52000	100	105000	1	
BARIUM	0	25.7	.4	66.7	50	92500	100	184000	.999984	
CADMIUM	0	4.30	.4	21.6	50	97200	100	190000	.999996	
CHROMIUM	0	6680	.4	6770	50	353000	100	705000	.999996	
COBALT	0	226	.4	269	50	306000	100	609000	.999959	
COPPER	0	110	.4	106	50	84100	100	171000	1	
LEAD	0	774	.4	923	50	894000	100	1760000	.999728	
MANGANESE	0	812	.4	892	50	312000	100	615000	.999906	
NICKEL	0	364	.4	348	50	84600	100	171000	1	
SILVER	0	71.3	.4	172	50	300000	100	609000	.99988	
THALLIUM	0	9.70	.4	134	50	381000	100	759000	.999873	
VANADIUM	0	1280	.4	1370	50	354000	100	719000	.999998	
ZINC	0	186	.4	165	50	50700	100	103000	.999997	

INT = Instrument intensity
R = Coefficient of correlation
Q = Data Qualifier
*** = Out of Compliance; R < 0.995**

INT_CAL_ICP - Modified 03/06/2008
PDF File ID: 4832323
Report generated: 07-JUL-2016 08:33

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16061026
Instrument ID: ICP-MS2
File ID: NI.062916.114146
Workgroup (AAB#): WG574572
Matrix: WATER

Run Date: 06/29/2016
Run Time: 11:41
Analyst: BKT
Cal ID: ICP-MS2 - 29-JUN-16

Sample ID: WG574676-06
Method: 6020A
Units: ug/L

Analytes	MDL	RDL	Concentration	Qualifier
SILVER	.2	.8	.2	U
ARSENIC	.2	.8	.2	U
BARIUM	.6	2.4	.6	U
CADMIUM	.12	.48	.12	U
COBALT	.2	.8	.2	U
CHROMIUM	.4	1.6	.4	U
COPPER	.4	1.6	.4	U
MANGANESE	.4	1.6	.4	U
NICKEL	.8	3.2	.8	U
LEAD	.2	.8	.2	U
ANTIMONY	.2	.8	.335	F
THALLIUM	.04	.16	.04	U
VANADIUM	.2	.8	.2	U
ZINC	5	20	5	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4832325
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16061026
Instrument ID: ICP-MS2
File ID: NI.070516.095329
Workgroup (AAB#): WG574572
Matrix: WATER

Run Date: 07/05/2016
Run Time: 09:53
Analyst: JYH
Cal ID: ICP-MS2 - 05-JUL-16

Sample ID: WG575179-09
Method: 6020A
Units: ug/L

Analytes	MDL	RDL	Concentration	Qualifier
SILVER	.2	.8	.2	U
ARSENIC	.2	.8	.2	U
BARIUM	.6	2.4	.6	U
CADMIUM	.12	.48	.12	U
COBALT	.2	.8	.2	U
CHROMIUM	.4	1.6	.4	U
COPPER	.4	1.6	.4	U
MANGANESE	.4	1.6	.4	U
NICKEL	.8	3.2	.8	U
LEAD	.2	.8	.2	U
ANTIMONY	.2	.8	.2	U
THALLIUM	.04	.16	.04	U
VANADIUM	.2	.8	.2	U
ZINC	5	20	5	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4832325
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
INITIAL CALIBRATION BLANK (ICB)

Login Number: L16061026
Instrument ID: ICP-MS2
File ID: NI.070616.092404
Workgroup (AAB#): WG574572
Matrix: WATER

Run Date: 07/06/2016
Run Time: 09:24
Analyst: JYH
Cal ID: ICP-MS2 - 06-JUL-16

Sample ID: WG575370-06
Method: 6020A
Units: ug/L

Analytes	MDL	RDL	Concentration	Qualifier
SILVER	.2	.8	.2	U
ARSENIC	.2	.8	.2	U
BARIUM	.6	2.4	.6	U
CADMIUM	.12	.48	.12	U
COBALT	.2	.8	.2	U
CHROMIUM	.4	1.6	.4	U
COPPER	.4	1.6	.4	U
MANGANESE	.4	1.6	.4	U
NICKEL	.8	3.2	.8	U
LEAD	.2	.8	.2	U
ANTIMONY	.2	.8	.236	F
THALLIUM	.04	.16	.0496	F
VANADIUM	.2	.8	.2	U
ZINC	5	20	5	U

U = Result is less than 2 x MDL

F = Result is between MDL and 2 x MDL

* = Result is above 2 x MDL

ICB - Modified 07/14/2009
PDF File ID: 4832325
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-11
Instrument ID: ICP-MS2
Run Time: 12:07
Method: 6020A
File ID: NI.062916.120734
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-13
Instrument ID: ICP-MS2
Run Time: 12:40
Method: 6020A
File ID: NI.062916.124059
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-16
Instrument ID: ICP-MS2
Run Time: 13:29
Method: 6020A
File ID: NI.062916.132901
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.834	F
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-18
Instrument ID: ICP-MS2
Run Time: 14:07
Method: 6020A
File ID: NI.062916.140723
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-20
Instrument ID: ICP-MS2
Run Time: 15:07
Method: 6020A
File ID: NI.062916.150718
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-28
Instrument ID: ICP-MS2
Run Time: 17:37
Method: 6020A
File ID: NI.062916.173730
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-32
Instrument ID: ICP-MS2
Run Time: 17:59
Method: 6020A
File ID: NI.062916.175955
Analyst: BKT
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.439	F
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-14
Instrument ID: ICP-MS2
Run Time: 10:09
Method: 6020A
File ID: NI.070516.100941
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.322	F
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-35
Instrument ID: ICP-MS2
Run Time: 15:44
Method: 6020A
File ID: NI.070516.154411
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-37
Instrument ID: ICP-MS2
Run Time: 16:21
Method: 6020A
File ID: NI.070516.162132
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-39
Instrument ID: ICP-MS2
Run Time: 16:50
Method: 6020A
File ID: NI.070516.165039
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.217	F
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-45
Instrument ID: ICP-MS2
Run Time: 18:39
Method: 6020A
File ID: NI.070516.183919
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-49
Instrument ID: ICP-MS2
Run Time: 19:17
Method: 6020A
File ID: NI.070516.191740
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0400	U
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-11
Instrument ID: ICP-MS2
Run Time: 09:40
Method: 6020A
File ID: NI.070616.094005
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.361	F
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0461	F
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-27
Instrument ID: ICP-MS2
Run Time: 14:55
Method: 6020A
File ID: NI.070616.145531
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.418	F
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0458	F
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-29
Instrument ID: ICP-MS2
Run Time: 15:34
Method: 6020A
File ID: NI.070616.153417
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0477	F
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-39
Instrument ID: ICP-MS2
Run Time: 18:53
Method: 6020A
File ID: NI.070616.185304
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.200	U
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0424	F
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION BLANK (CCB)

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-43
Instrument ID: ICP-MS2
Run Time: 19:15
Method: 6020A
File ID: NI.070616.191526
Analyst: JYH
Units: ug/L
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER
QAPP: DOD4

Analytes	MDL	RDL	Concentration	Qualifier
Antimony	0.200	0.800	0.298	F
Arsenic	0.200	0.800	0.200	U
Barium	0.600	2.40	0.600	U
Cadmium	0.120	0.480	0.120	U
Chromium	0.400	1.60	0.400	U
Cobalt	0.200	0.800	0.200	U
Copper	0.400	1.60	0.400	U
Lead	0.200	0.800	0.200	U
Manganese	0.400	1.60	0.400	U
Nickel	0.800	3.20	0.800	U
Silver	0.200	0.800	0.200	U
Thallium	0.0400	0.160	0.0490	F
Vanadium	0.200	0.800	0.200	U
Zinc	5.00	20.0	5.00	U

U = Result is less than MDL.
F = Result is between MDL and RL.
* = Result is above RL.

CCB - Modified 03/05/2008
PDF File ID: 4832328
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16061026 Run Date: 07/06/2016 Sample ID: WG575370-05
Instrument ID: ICP-MS2 Run Time: 09:20 Method: 6020A
File ID: NI.070616.092051 Analyst: JYH Units: ug/L
Workgroup (AAB#): WG574572 Cal ID: ICP-MS - 06-JUL-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Antimony		50	51.8	104	90 - 110	
Arsenic		50	50.2	100	90 - 110	
Barium		50	51.1	102	90 - 110	
Cadmium		50	50.8	102	90 - 110	
Chromium		50	50.2	100	90 - 110	
Cobalt		50	50.7	101	90 - 110	
Copper		50	50.0	100	90 - 110	
Lead		50	50.8	102	90 - 110	
Manganese		50	50.7	101	90 - 110	
Nickel		50	49.7	99.4	90 - 110	
Silver		50	50.9	102	90 - 110	
Thallium		50	50.3	101	90 - 110	
Vanadium		50	50.7	101	90 - 110	
Zinc		50	50.3	101	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4832324
Report generated 07/07/2016 08:33

Microbac Laboratories Inc.

INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16061026 Run Date: 06/29/2016 Sample ID: WG574676-05
Instrument ID: ICP-MS2 Run Time: 11:38 Method: 6020A
File ID: NI.062916.113833 Analyst: BKT Units: ug/L
Workgroup (AAB#): WG574572 Cal ID: ICP-MS - 29-JUN-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Antimony		50	49.8	99.7	90 - 110	
Arsenic		50	50.9	102	90 - 110	
Barium		50	49.9	99.9	90 - 110	
Cadmium		50	50.3	101	90 - 110	
Chromium		50	50.0	100	90 - 110	
Cobalt		50	50.4	101	90 - 110	
Copper		50	50.8	102	90 - 110	
Lead		50	51.0	102	90 - 110	
Manganese		50	51.9	104	90 - 110	
Nickel		50	50.6	101	90 - 110	
Silver		50	50.9	102	90 - 110	
Thallium		50	50.7	101	90 - 110	
Vanadium		50	51.2	102	90 - 110	
Zinc		50	52.1	104	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4832324
Report generated 07/07/2016 08:33

Microbac Laboratories Inc.
INITIAL CALIBRATION VERIFICATION (ICV)
(Alternate Source)

Login Number: L16061026 Run Date: 07/05/2016 Sample ID: WG575179-08
Instrument ID: ICP-MS2 Run Time: 09:50 Method: 6020A
File ID: NI.070516.095016 Analyst: JYH Units: ug/L
Workgroup (AAB#): WG574572 Cal ID: ICP-MS - 05-JUL-16
QC Key: DOD4

Analyte		Expected	Found	%REC	LIMITS	Q
Antimony		50	49.1	98.2	90 - 110	
Arsenic		50	47.5	95.1	90 - 110	
Barium		50	47.9	95.8	90 - 110	
Cadmium		50	47.9	95.8	90 - 110	
Chromium		50	48.0	96.0	90 - 110	
Cobalt		50	48.4	96.8	90 - 110	
Copper		50	48.0	96.1	90 - 110	
Lead		50	48.1	96.1	90 - 110	
Manganese		50	48.1	96.3	90 - 110	
Nickel		50	48.0	96.0	90 - 110	
Silver		50	48.0	95.9	90 - 110	
Thallium		50	48.0	96.0	90 - 110	
Vanadium		50	48.2	96.3	90 - 110	
Zinc		50	48.3	96.6	90 - 110	

* Exceeds LIMITS Limit

ICV - Modified 03/06/2008
PDF File ID: 4832324
Report generated 07/07/2016 08:33

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-10
Instrument ID: ICP-MS2
Run Time: 12:00
Method: 6020A
File ID: NI.062916.120010
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0508	mg/L	102	90 - 110		
Arsenic		0.0500	0.0514	mg/L	103	90 - 110		
Barium		0.0500	0.0505	mg/L	101	90 - 110		
Cadmium		0.0500	0.0516	mg/L	103	90 - 110		
Chromium		0.0500	0.0504	mg/L	101	90 - 110		
Cobalt		0.0500	0.0538	mg/L	108	90 - 110		
Copper		0.0500	0.0507	mg/L	101	90 - 110		
Lead		0.0500	0.0529	mg/L	106	90 - 110		
Manganese		0.0500	0.0534	mg/L	107	90 - 110		
Nickel		0.0500	0.0502	mg/L	100	90 - 110		
Silver		0.0500	0.0508	mg/L	102	90 - 110		
Thallium		0.0500	0.0519	mg/L	104	90 - 110		
Vanadium		0.0500	0.0516	mg/L	103	90 - 110		
Zinc		0.0500	0.0510	mg/L	102	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-12
Instrument ID: ICP-MS2
Run Time: 12:37
Method: 6020A
File ID: NI.062916.123748
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0505	mg/L	101	90 - 110		
Arsenic		0.0500	0.0498	mg/L	99.6	90 - 110		
Barium		0.0500	0.0499	mg/L	99.8	90 - 110		
Cadmium		0.0500	0.0506	mg/L	101	90 - 110		
Chromium		0.0500	0.0489	mg/L	97.8	90 - 110		
Cobalt		0.0500	0.0506	mg/L	101	90 - 110		
Copper		0.0500	0.0500	mg/L	99.9	90 - 110		
Lead		0.0500	0.0499	mg/L	99.7	90 - 110		
Manganese		0.0500	0.0500	mg/L	99.9	90 - 110		
Nickel		0.0500	0.0491	mg/L	98.2	90 - 110		
Silver		0.0500	0.0505	mg/L	101	90 - 110		
Thallium		0.0500	0.0497	mg/L	99.4	90 - 110		
Vanadium		0.0500	0.0503	mg/L	101	90 - 110		
Zinc		0.0500	0.0498	mg/L	99.7	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-15
Instrument ID: ICP-MS2
Run Time: 13:25
Method: 6020A
File ID: NI.062916.132550
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0490	mg/L	98.1	90 - 110		
Arsenic		0.0500	0.0501	mg/L	100	90 - 110		
Barium		0.0500	0.0473	mg/L	94.6	90 - 110		
Cadmium		0.0500	0.0490	mg/L	98.0	90 - 110		
Chromium		0.0500	0.0489	mg/L	97.9	90 - 110		
Cobalt		0.0500	0.0557	mg/L	111	90 - 110		*
Copper		0.0500	0.0491	mg/L	98.3	90 - 110		
Lead		0.0500	0.0505	mg/L	101	90 - 110		
Manganese		0.0500	0.0508	mg/L	102	90 - 110		
Nickel		0.0500	0.0487	mg/L	97.5	90 - 110		
Silver		0.0500	0.0499	mg/L	99.7	90 - 110		
Thallium		0.0500	0.0494	mg/L	98.8	90 - 110		
Vanadium		0.0500	0.0520	mg/L	104	90 - 110		
Zinc		0.0500	0.0498	mg/L	99.6	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026 Run Date: 06/29/2016 Sample ID: WG574676-17
Instrument ID: ICP-MS2 Run Time: 14:04 Method: 6020A
File ID: NI.062916.140412 Analyst: BKT QC Key: DOD4
Workgroup (AAB#): WG574572 Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0501	mg/L	100	90 - 110		
Arsenic		0.0500	0.0523	mg/L	105	90 - 110		
Barium		0.0500	0.0474	mg/L	94.8	90 - 110		
Cadmium		0.0500	0.0492	mg/L	98.4	90 - 110		
Chromium		0.0500	0.0510	mg/L	102	90 - 110		
Cobalt		0.0500	0.0568	mg/L	114	90 - 110		*
Copper		0.0500	0.0515	mg/L	103	90 - 110		
Lead		0.0500	0.0511	mg/L	102	90 - 110		
Manganese		0.0500	0.0517	mg/L	103	90 - 110		
Nickel		0.0500	0.0509	mg/L	102	90 - 110		
Silver		0.0500	0.0508	mg/L	102	90 - 110		
Thallium		0.0500	0.0506	mg/L	101	90 - 110		
Vanadium		0.0500	0.0539	mg/L	108	90 - 110		
Zinc		0.0500	0.0522	mg/L	104	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026 Run Date: 06/29/2016 Sample ID: WG574676-19
Instrument ID: ICP-MS2 Run Time: 15:04 Method: 6020A
File ID: NI.062916.150407 Analyst: BKT QC Key: DOD4
Workgroup (AAB#): WG574572 Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0495	mg/L	98.9	90 - 110		
Arsenic		0.0500	0.0504	mg/L	101	90 - 110		
Barium		0.0500	0.0484	mg/L	96.9	90 - 110		
Cadmium		0.0500	0.0498	mg/L	99.6	90 - 110		
Chromium		0.0500	0.0505	mg/L	101	90 - 110		
Cobalt		0.0500	0.0552	mg/L	110	90 - 110		
Copper		0.0500	0.0505	mg/L	101	90 - 110		
Lead		0.0500	0.0526	mg/L	105	90 - 110		
Manganese		0.0500	0.0531	mg/L	106	90 - 110		
Nickel		0.0500	0.0503	mg/L	101	90 - 110		
Silver		0.0500	0.0501	mg/L	100	90 - 110		
Thallium		0.0500	0.0509	mg/L	102	90 - 110		
Vanadium		0.0500	0.0523	mg/L	105	90 - 110		
Zinc		0.0500	0.0512	mg/L	102	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-27
Instrument ID: ICP-MS2
Run Time: 17:34
Method: 6020A
File ID: NI.062916.173419
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0486	mg/L	97.2	90 - 110		
Arsenic		0.0500	0.0488	mg/L	97.6	90 - 110		
Barium		0.0500	0.0466	mg/L	93.2	90 - 110		
Cadmium		0.0500	0.0485	mg/L	97.1	90 - 110		
Chromium		0.0500	0.0491	mg/L	98.2	90 - 110		
Cobalt		0.0500	0.0513	mg/L	103	90 - 110		
Copper		0.0500	0.0491	mg/L	98.1	90 - 110		
Lead		0.0500	0.0503	mg/L	101	90 - 110		
Manganese		0.0500	0.0482	mg/L	96.4	90 - 110		
Nickel		0.0500	0.0489	mg/L	97.8	90 - 110		
Silver		0.0500	0.0493	mg/L	98.6	90 - 110		
Thallium		0.0500	0.0491	mg/L	98.2	90 - 110		
Vanadium		0.0500	0.0503	mg/L	101	90 - 110		
Zinc		0.0500	0.0491	mg/L	98.1	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-31
Instrument ID: ICP-MS2
Run Time: 17:56
Method: 6020A
File ID: NI.062916.175644
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0502	mg/L	100	90 - 110		
Arsenic		0.0500	0.0511	mg/L	102	90 - 110		
Barium		0.0500	0.0484	mg/L	96.9	90 - 110		
Cadmium		0.0500	0.0504	mg/L	101	90 - 110		
Chromium		0.0500	0.0524	mg/L	105	90 - 110		
Cobalt		0.0500	0.0556	mg/L	111	90 - 110		*
Copper		0.0500	0.0514	mg/L	103	90 - 110		
Lead		0.0500	0.0535	mg/L	107	90 - 110		
Manganese		0.0500	0.0515	mg/L	103	90 - 110		
Nickel		0.0500	0.0517	mg/L	103	90 - 110		
Silver		0.0500	0.0504	mg/L	101	90 - 110		
Thallium		0.0500	0.0525	mg/L	105	90 - 110		
Vanadium		0.0500	0.0529	mg/L	106	90 - 110		
Zinc		0.0500	0.0512	mg/L	102	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-13
Instrument ID: ICP-MS2
Run Time: 10:06
Method: 6020A
File ID: NI.070516.100630
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0520	mg/L	104	90 - 110		
Arsenic		0.0500	0.0487	mg/L	97.3	90 - 110		
Barium		0.0500	0.0490	mg/L	98.0	90 - 110		
Cadmium		0.0500	0.0495	mg/L	99.1	90 - 110		
Chromium		0.0500	0.0494	mg/L	98.9	90 - 110		
Cobalt		0.0500	0.0498	mg/L	99.6	90 - 110		
Copper		0.0500	0.0491	mg/L	98.3	90 - 110		
Lead		0.0500	0.0502	mg/L	100	90 - 110		
Manganese		0.0500	0.0491	mg/L	98.3	90 - 110		
Nickel		0.0500	0.0496	mg/L	99.2	90 - 110		
Silver		0.0500	0.0494	mg/L	98.8	90 - 110		
Thallium		0.0500	0.0500	mg/L	100	90 - 110		
Vanadium		0.0500	0.0498	mg/L	99.6	90 - 110		
Zinc		0.0500	0.0497	mg/L	99.4	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-34
Instrument ID: ICP-MS2
Run Time: 15:41
Method: 6020A
File ID: NI.070516.154100
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0493	mg/L	98.6	90 - 110		
Arsenic		0.0500	0.0472	mg/L	94.5	90 - 110		
Barium		0.0500	0.0462	mg/L	92.3	90 - 110		
Cadmium		0.0500	0.0473	mg/L	94.6	90 - 110		
Chromium		0.0500	0.0464	mg/L	92.8	90 - 110		
Cobalt		0.0500	0.0480	mg/L	96.0	90 - 110		
Copper		0.0500	0.0471	mg/L	94.2	90 - 110		
Lead		0.0500	0.0499	mg/L	99.7	90 - 110		
Manganese		0.0500	0.0452	mg/L	90.4	90 - 110		
Nickel		0.0500	0.0467	mg/L	93.3	90 - 110		
Silver		0.0500	0.0474	mg/L	94.8	90 - 110		
Thallium		0.0500	0.0478	mg/L	95.5	90 - 110		
Vanadium		0.0500	0.0465	mg/L	93.0	90 - 110		
Zinc		0.0500	0.0502	mg/L	100	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-36
Instrument ID: ICP-MS2
Run Time: 16:18
Method: 6020A
File ID: NI.070516.161820
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0477	mg/L	95.4	90 - 110		
Arsenic		0.0500	0.0479	mg/L	95.8	90 - 110		
Barium		0.0500	0.0456	mg/L	91.2	90 - 110		
Cadmium		0.0500	0.0471	mg/L	94.3	90 - 110		
Chromium		0.0500	0.0473	mg/L	94.7	90 - 110		
Cobalt		0.0500	0.0491	mg/L	98.2	90 - 110		
Copper		0.0500	0.0478	mg/L	95.5	90 - 110		
Lead		0.0500	0.0505	mg/L	101	90 - 110		
Manganese		0.0500	0.0463	mg/L	92.7	90 - 110		
Nickel		0.0500	0.0472	mg/L	94.3	90 - 110		
Silver		0.0500	0.0467	mg/L	93.4	90 - 110		
Thallium		0.0500	0.0479	mg/L	95.7	90 - 110		
Vanadium		0.0500	0.0470	mg/L	94.1	90 - 110		
Zinc		0.0500	0.0491	mg/L	98.1	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-38
Instrument ID: ICP-MS2
Run Time: 16:47
Method: 6020A
File ID: NI.070516.164727
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0516	mg/L	103	90 - 110		
Arsenic		0.0500	0.0499	mg/L	99.8	90 - 110		
Barium		0.0500	0.0480	mg/L	95.9	90 - 110		
Cadmium		0.0500	0.0496	mg/L	99.3	90 - 110		
Chromium		0.0500	0.0492	mg/L	98.4	90 - 110		
Cobalt		0.0500	0.0503	mg/L	101	90 - 110		
Copper		0.0500	0.0492	mg/L	98.3	90 - 110		
Lead		0.0500	0.0521	mg/L	104	90 - 110		
Manganese		0.0500	0.0483	mg/L	96.6	90 - 110		
Nickel		0.0500	0.0490	mg/L	98.0	90 - 110		
Silver		0.0500	0.0492	mg/L	98.5	90 - 110		
Thallium		0.0500	0.0500	mg/L	100	90 - 110		
Vanadium		0.0500	0.0488	mg/L	97.6	90 - 110		
Zinc		0.0500	0.0510	mg/L	102	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-44
Instrument ID: ICP-MS2
Run Time: 18:36
Method: 6020A
File ID: NI.070516.183607
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0534	mg/L	107	90 - 110		
Arsenic		0.0500	0.0512	mg/L	102	90 - 110		
Barium		0.0500	0.0500	mg/L	100	90 - 110		
Cadmium		0.0500	0.0514	mg/L	103	90 - 110		
Chromium		0.0500	0.0508	mg/L	102	90 - 110		
Cobalt		0.0500	0.0523	mg/L	105	90 - 110		
Copper		0.0500	0.0510	mg/L	102	90 - 110		
Lead		0.0500	0.0546	mg/L	109	90 - 110		
Manganese		0.0500	0.0500	mg/L	100	90 - 110		
Nickel		0.0500	0.0513	mg/L	103	90 - 110		
Silver		0.0500	0.0508	mg/L	102	90 - 110		
Thallium		0.0500	0.0521	mg/L	104	90 - 110		
Vanadium		0.0500	0.0503	mg/L	101	90 - 110		
Zinc		0.0500	0.0529	mg/L	106	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-48
Instrument ID: ICP-MS2
Run Time: 19:14
Method: 6020A
File ID: NI.070516.191429
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0500	mg/L	100	90 - 110		
Arsenic		0.0500	0.0478	mg/L	95.5	90 - 110		
Barium		0.0500	0.0462	mg/L	92.3	90 - 110		
Cadmium		0.0500	0.0476	mg/L	95.3	90 - 110		
Chromium		0.0500	0.0470	mg/L	94.0	90 - 110		
Cobalt		0.0500	0.0486	mg/L	97.3	90 - 110		
Copper		0.0500	0.0475	mg/L	94.9	90 - 110		
Lead		0.0500	0.0507	mg/L	101	90 - 110		
Manganese		0.0500	0.0465	mg/L	92.9	90 - 110		
Nickel		0.0500	0.0472	mg/L	94.3	90 - 110		
Silver		0.0500	0.0472	mg/L	94.3	90 - 110		
Thallium		0.0500	0.0484	mg/L	96.8	90 - 110		
Vanadium		0.0500	0.0472	mg/L	94.4	90 - 110		
Zinc		0.0500	0.0491	mg/L	98.2	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-10
Instrument ID: ICP-MS2
Run Time: 09:36
Method: 6020A
File ID: NI.070616.093653
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0511	mg/L	102	90 - 110		
Arsenic		0.0500	0.0505	mg/L	101	90 - 110		
Barium		0.0500	0.0490	mg/L	97.9	90 - 110		
Cadmium		0.0500	0.0493	mg/L	98.7	90 - 110		
Chromium		0.0500	0.0505	mg/L	101	90 - 110		
Cobalt		0.0500	0.0509	mg/L	102	90 - 110		
Copper		0.0500	0.0504	mg/L	101	90 - 110		
Lead		0.0500	0.0511	mg/L	102	90 - 110		
Manganese		0.0500	0.0513	mg/L	103	90 - 110		
Nickel		0.0500	0.0502	mg/L	100	90 - 110		
Silver		0.0500	0.0493	mg/L	98.6	90 - 110		
Thallium		0.0500	0.0505	mg/L	101	90 - 110		
Vanadium		0.0500	0.0507	mg/L	101	90 - 110		
Zinc		0.0500	0.0507	mg/L	101	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-26
Instrument ID: ICP-MS2
Run Time: 14:52
Method: 6020A
File ID: NI.070616.145220
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0512	mg/L	102	90 - 110		
Arsenic		0.0500	0.0508	mg/L	102	90 - 110		
Barium		0.0500	0.0499	mg/L	99.8	90 - 110		
Cadmium		0.0500	0.0508	mg/L	102	90 - 110		
Chromium		0.0500	0.0519	mg/L	104	90 - 110		
Cobalt		0.0500	0.0506	mg/L	101	90 - 110		
Copper		0.0500	0.0515	mg/L	103	90 - 110		
Lead		0.0500	0.0525	mg/L	105	90 - 110		
Manganese		0.0500	0.0530	mg/L	106	90 - 110		
Nickel		0.0500	0.0518	mg/L	104	90 - 110		
Silver		0.0500	0.0499	mg/L	99.7	90 - 110		
Thallium		0.0500	0.0508	mg/L	102	90 - 110		
Vanadium		0.0500	0.0504	mg/L	101	90 - 110		
Zinc		0.0500	0.0512	mg/L	102	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-28
Instrument ID: ICP-MS2
Run Time: 15:31
Method: 6020A
File ID: NI.070616.153105
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0504	mg/L	101	90 - 110		
Arsenic		0.0500	0.0512	mg/L	102	90 - 110		
Barium		0.0500	0.0495	mg/L	99.0	90 - 110		
Cadmium		0.0500	0.0505	mg/L	101	90 - 110		
Chromium		0.0500	0.0523	mg/L	105	90 - 110		
Cobalt		0.0500	0.0520	mg/L	104	90 - 110		
Copper		0.0500	0.0523	mg/L	105	90 - 110		
Lead		0.0500	0.0529	mg/L	106	90 - 110		
Manganese		0.0500	0.0537	mg/L	107	90 - 110		
Nickel		0.0500	0.0525	mg/L	105	90 - 110		
Silver		0.0500	0.0502	mg/L	100	90 - 110		
Thallium		0.0500	0.0513	mg/L	103	90 - 110		
Vanadium		0.0500	0.0512	mg/L	102	90 - 110		
Zinc		0.0500	0.0519	mg/L	104	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-38
Instrument ID: ICP-MS2
Run Time: 18:49
Method: 6020A
File ID: NI.070616.184952
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0496	mg/L	99.2	90 - 110		
Arsenic		0.0500	0.0503	mg/L	101	90 - 110		
Barium		0.0500	0.0491	mg/L	98.3	90 - 110		
Cadmium		0.0500	0.0501	mg/L	100	90 - 110		
Chromium		0.0500	0.0517	mg/L	103	90 - 110		
Cobalt		0.0500	0.0499	mg/L	99.8	90 - 110		
Copper		0.0500	0.0514	mg/L	103	90 - 110		
Lead		0.0500	0.0522	mg/L	104	90 - 110		
Manganese		0.0500	0.0527	mg/L	105	90 - 110		
Nickel		0.0500	0.0516	mg/L	103	90 - 110		
Silver		0.0500	0.0490	mg/L	98.0	90 - 110		
Thallium		0.0500	0.0508	mg/L	102	90 - 110		
Vanadium		0.0500	0.0496	mg/L	99.2	90 - 110		
Zinc		0.0500	0.0512	mg/L	102	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

CONTINUING CALIBRATION VERIFICATION (CCV)

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-42
Instrument ID: ICP-MS2
Run Time: 19:12
Method: 6020A
File ID: NI.070616.191215
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.0500	0.0526	mg/L	105	90 - 110		
Arsenic		0.0500	0.0506	mg/L	101	90 - 110		
Barium		0.0500	0.0495	mg/L	99.0	90 - 110		
Cadmium		0.0500	0.0507	mg/L	101	90 - 110		
Chromium		0.0500	0.0517	mg/L	103	90 - 110		
Cobalt		0.0500	0.0513	mg/L	103	90 - 110		
Copper		0.0500	0.0513	mg/L	103	90 - 110		
Lead		0.0500	0.0521	mg/L	104	90 - 110		
Manganese		0.0500	0.0530	mg/L	106	90 - 110		
Nickel		0.0500	0.0514	mg/L	103	90 - 110		
Silver		0.0500	0.0501	mg/L	100	90 - 110		
Thallium		0.0500	0.0512	mg/L	102	90 - 110		
Vanadium		0.0500	0.0506	mg/L	101	90 - 110		
Zinc		0.0500	0.0513	mg/L	103	90 - 110		

* Exceeds LIMITS Criteria

CCV - Modified 03/05/2008
PDF File ID: 4832327
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-07
Instrument ID: ICP-MS2
Run Time: 11:46
Method: 6020A
File ID: NI.062916.114653
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.456	ug/L	114	70 - 130		
Arsenic		0.400	0.429	ug/L	107	70 - 130		
Barium		0.750	0.733	ug/L	97.8	70 - 130		
Cadmium		0.240	0.237	ug/L	98.5	70 - 130		
Chromium		0.800	0.739	ug/L	92.4	70 - 130		
Cobalt		0.400	0.414	ug/L	103	70 - 130		
Copper		0.800	0.768	ug/L	96.0	70 - 130		
Lead		0.200	0.181	ug/L	90.3	70 - 130		
Manganese		0.500	0.468	ug/L	93.6	70 - 130		
Nickel		1.60	1.56	ug/L	97.7	70 - 130		
Silver		0.400	0.390	ug/L	97.5	70 - 130		
Thallium		0.0800	0.0871	ug/L	109	70 - 130		
Vanadium		0.400	0.382	ug/L	95.5	70 - 130		
Zinc		6.25	6.64	ug/L	106	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4832329
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-14
Instrument ID: ICP-MS2
Run Time: 12:44
Method: 6020A
File ID: NI.062916.124412
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.401	ug/L	100	70 - 130		
Arsenic		0.400	0.394	ug/L	98.6	70 - 130		
Barium		0.750	0.718	ug/L	95.7	70 - 130		
Cadmium		0.240	0.227	ug/L	94.7	70 - 130		
Chromium		0.800	0.768	ug/L	96.0	70 - 130		
Cobalt		0.400	0.405	ug/L	101	70 - 130		
Copper		0.800	0.751	ug/L	93.9	70 - 130		
Lead		0.200	0.179	ug/L	89.4	70 - 130		
Manganese		0.500	0.469	ug/L	93.8	70 - 130		
Nickel		1.60	1.53	ug/L	95.6	70 - 130		
Silver		0.400	0.383	ug/L	95.7	70 - 130		
Thallium		0.0800	0.0809	ug/L	101	70 - 130		
Vanadium		0.400	0.370	ug/L	92.4	70 - 130		
Zinc		6.25	6.89	ug/L	110	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4832329
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16061026
Run Date: 06/29/2016
Sample ID: WG574676-33
Instrument ID: ICP-MS2
Run Time: 18:03
Method: 6020A
File ID: NI.062916.180308
Analyst: BKT
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 29-JUN-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.445	ug/L	111	70 - 130		
Arsenic		0.400	0.497	ug/L	124	70 - 130		
Barium		0.750	0.689	ug/L	91.9	70 - 130		
Cadmium		0.240	0.226	ug/L	94.3	70 - 130		
Chromium		0.800	0.764	ug/L	95.5	70 - 130		
Cobalt		0.400	0.420	ug/L	105	70 - 130		
Copper		0.800	0.807	ug/L	101	70 - 130		
Lead		0.200	0.187	ug/L	93.5	70 - 130		
Manganese		0.500	0.444	ug/L	88.8	70 - 130		
Nickel		1.60	1.62	ug/L	101	70 - 130		
Silver		0.400	0.401	ug/L	100	70 - 130		
Thallium		0.0800	0.0936	ug/L	117	70 - 130		
Vanadium		0.400	0.389	ug/L	97.4	70 - 130		
Zinc		6.25	6.56	ug/L	105	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4832329
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-10
Instrument ID: ICP-MS2
Run Time: 09:56
Method: 6020A
File ID: NI.070516.095642
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.365	ug/L	91.1	70 - 130		
Arsenic		0.400	0.341	ug/L	85.2	70 - 130		
Barium		0.750	0.667	ug/L	88.9	70 - 130		
Cadmium		0.240	0.221	ug/L	92.2	70 - 130		
Chromium		0.800	0.612	ug/L	76.5	70 - 130		
Cobalt		0.400	0.364	ug/L	90.9	70 - 130		
Copper		0.800	0.720	ug/L	90.0	70 - 130		
Lead		0.200	0.178	ug/L	88.8	70 - 130		
Manganese		0.500	0.427	ug/L	85.4	70 - 130		
Nickel		1.60	1.39	ug/L	87.1	70 - 130		
Silver		0.400	0.360	ug/L	90.0	70 - 130		
Thallium		0.0800	0.0696	ug/L	87.0	70 - 130		
Vanadium		0.400	0.325	ug/L	81.2	70 - 130		
Zinc		6.25	6.36	ug/L	102	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4832329
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-31
Instrument ID: ICP-MS2
Run Time: 14:36
Method: 6020A
File ID: NI.070516.143603
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.367	ug/L	91.8	70 - 130		
Arsenic		0.400	0.475	ug/L	119	70 - 130		
Barium		0.750	0.629	ug/L	83.8	70 - 130		
Cadmium		0.240	0.213	ug/L	88.6	70 - 130		
Chromium		0.800	0.664	ug/L	83.0	70 - 130		
Cobalt		0.400	0.365	ug/L	91.3	70 - 130		
Copper		0.800	0.937	ug/L	117	70 - 130		
Lead		0.200	0.182	ug/L	91.0	70 - 130		
Manganese		0.500	0.362	ug/L	72.4	70 - 130		
Nickel		1.60	1.44	ug/L	89.8	70 - 130		
Silver		0.400	0.358	ug/L	89.4	70 - 130		
Thallium		0.0800	0.0750	ug/L	93.8	70 - 130		
Vanadium		0.400	0.396	ug/L	99.0	70 - 130		
Zinc		6.25	6.51	ug/L	104	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4832329
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16061026
Run Date: 07/05/2016
Sample ID: WG575179-50
Instrument ID: ICP-MS2
Run Time: 19:20
Method: 6020A
File ID: NI.070516.192053
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 05-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.370	ug/L	92.5	70 - 130		
Arsenic		0.400	0.371	ug/L	92.7	70 - 130		
Barium		0.750	0.628	ug/L	83.7	70 - 130		
Cadmium		0.240	0.215	ug/L	89.5	70 - 130		
Chromium		0.800	0.520	ug/L	65.0	70 - 130		*
Cobalt		0.400	0.363	ug/L	90.7	70 - 130		
Copper		0.800	0.679	ug/L	84.9	70 - 130		
Lead		0.200	0.185	ug/L	92.4	70 - 130		
Manganese		0.500	0.423	ug/L	84.6	70 - 130		
Nickel		1.60	1.32	ug/L	82.4	70 - 130		
Silver		0.400	0.342	ug/L	85.4	70 - 130		
Thallium		0.0800	0.0766	ug/L	95.8	70 - 130		
Vanadium		0.400	0.317	ug/L	79.3	70 - 130		
Zinc		6.25	5.99	ug/L	95.9	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4832329
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-07
Instrument ID: ICP-MS2
Run Time: 09:27
Method: 6020A
File ID: NI.070616.092717
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.435	ug/L	109	70 - 130		
Arsenic		0.400	0.418	ug/L	105	70 - 130		
Barium		0.750	0.739	ug/L	98.5	70 - 130		
Cadmium		0.240	0.274	ug/L	114	70 - 130		
Chromium		0.800	0.743	ug/L	92.9	70 - 130		
Cobalt		0.400	0.423	ug/L	106	70 - 130		
Copper		0.800	0.816	ug/L	102	70 - 130		
Lead		0.200	0.240	ug/L	120	70 - 130		
Manganese		0.500	0.495	ug/L	99.0	70 - 130		
Nickel		1.60	1.57	ug/L	98.4	70 - 130		
Silver		0.400	0.428	ug/L	107	70 - 130		
Thallium		0.0800	0.120	ug/L	150	70 - 130		*
Vanadium		0.400	0.398	ug/L	99.5	70 - 130		
Zinc		6.25	6.43	ug/L	103	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4832329
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
LOW LEVEL CALIBRATION VERIFICATION

Login Number: L16061026
Run Date: 07/06/2016
Sample ID: WG575370-44
Instrument ID: ICP-MS2
Run Time: 19:18
Method: 6020A
File ID: NI.070616.191839
Analyst: JYH
QC Key: DOD4
Workgroup (AAB#): WG574572
Cal ID: ICP-MS - 06-JUL-16
Matrix: WATER

Analyte		Expected	Found	UNITS	%REC	LIMITS		Q
Antimony		0.400	0.451	ug/L	113	70 - 130		
Arsenic		0.400	0.437	ug/L	109	70 - 130		
Barium		0.750	0.726	ug/L	96.8	70 - 130		
Cadmium		0.240	0.283	ug/L	118	70 - 130		
Chromium		0.800	0.809	ug/L	101	70 - 130		
Cobalt		0.400	0.419	ug/L	105	70 - 130		
Copper		0.800	0.837	ug/L	105	70 - 130		
Lead		0.200	0.243	ug/L	121	70 - 130		
Manganese		0.500	0.498	ug/L	99.7	70 - 130		
Nickel		1.60	1.57	ug/L	98.0	70 - 130		
Silver		0.400	0.418	ug/L	105	70 - 130		
Thallium		0.0800	0.113	ug/L	142	70 - 130		*
Vanadium		0.400	0.395	ug/L	98.8	70 - 130		
Zinc		6.25	6.69	ug/L	107	70 - 130		

* Exceeds LIMITS Criteria

LLCCV - Modified 1/7/2010
PDF File ID: 4832329
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16061026
Instrument ID: ICP-MS2
Sol. A : WG574676-08 **File ID:** NI.062916.115053
Sol. AB : WG574676-09 **File ID:** NI.062916.115447

Workgroup (AAB#): WG574572
Method: 6020A
Units: ug/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Antimony	NS	0.0658	NS	100	100	100	
Arsenic	NS	0.0161	NS	100	105	105	
Barium	NS	0.303	NS	100	98.4	98.4	
Cadmium	NS	-0.103	NS	100	101	101	
Chromium	NS	0.0680	NS	100	101	101	
Cobalt	NS	0.0292	NS	100	105	105	
Copper	NS	0.245	NS	100	100	100	
Lead	NS	0.0174	NS	100	102	102	
Manganese	NS	0.0629	NS	100	105	105	
Nickel	NS	0.331	NS	100	100	100	
Silver	NS	0.0271	NS	100	95.3	95.3	
Thallium	NS	0.0271	NS	100	100	100	
Vanadium	NS	0.0706	NS	100	103	103	
Zinc	NS	1.14	NS	100	104	104	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4832326
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16061026
Instrument ID: ICP-MS2
Sol. A : WG575179-11 **File ID:** NI.070516.095958
Sol. AB : WG575179-12 **File ID:** NI.070516.100317

Workgroup (AAB#): WG574572
Method: 6020A
Units: ug/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Antimony	NS	0.133	NS	100	96.1	96.1	
Arsenic	NS	-0.00770	NS	100	97.9	97.9	
Barium	NS	0.724	NS	100	94.5	94.5	
Cadmium	NS	-0.0970	NS	100	96.4	96.4	
Chromium	NS	0.0752	NS	100	94.6	94.6	
Cobalt	NS	0.0303	NS	100	94.9	94.9	
Copper	NS	0.309	NS	100	94.8	94.8	
Lead	NS	0.112	NS	100	96.0	96.0	
Manganese	NS	0.227	NS	100	96.2	96.2	
Nickel	NS	0.346	NS	100	94.6	94.6	
Silver	NS	0.0320	NS	100	83.4	83.4	
Thallium	NS	0.0119	NS	100	95.8	95.8	
Vanadium	NS	0.0220	NS	100	95.9	95.9	
Zinc	NS	1.07	NS	100	98.0	98.0	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4832326
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.
INTERFERENCE CHECK SAMPLES

Login number: L16061026
Instrument ID: ICP-MS2
Sol. A : WG575370-08 **File ID:** NI.070616.093029
Sol. AB : WG575370-09 **File ID:** NI.070616.093340

Workgroup (AAB#): WG574572
Method: 6020A
Units: ug/L
Matrix: Water

	Sol. A			Sol. AB			
ANALYTE	True	Found	%Recovery	True	Found	%Recovery	Q
Antimony	NS	0.462	NS	100	101	101	
Arsenic	NS	0.0530	NS	100	103	103	
Barium	NS	0.0414	NS	100	101	101	
Cadmium	NS	-0.0508	NS	100	101	101	
Chromium	NS	-0.0525	NS	100	101	101	
Cobalt	NS	0.0720	NS	100	101	101	
Copper	NS	0.326	NS	100	102	102	
Lead	NS	0.0662	NS	100	103	103	
Manganese	NS	0.305	NS	100	102	102	
Nickel	NS	0.313	NS	100	100	100	
Silver	NS	0.0592	NS	100	85.4	85.4	
Thallium	NS	0.0459	NS	100	101	101	
Vanadium	NS	0.0677	NS	100	102	102	
Zinc	NS	0.776	NS	100	103	103	

NS = Not spiked
*** = Recovery of spiked element is outside acceptance limit of 80% - 120% of true value.**
 # = Result for unspiked element is outside the acceptance limits of (+/-) the project reporting limit (RL).
+ = Result for unspiked element is outside the acceptance limits of (+/-) 2 times the project method detection limit (MDL). This criteria is only applicable to specific QAPPs.

ICS - Modified 03/06/2008
PDF File ID: 4832326
Report generated 07/07/2016 08:34

Microbac Laboratories Inc.

INTERNAL STANDARD REPORT

Login: L16061026 **Analytical Method:** 6020
Analytical Workgroup: WG574572 **Matrix:** 1
Instrument: ICP-MS2 **Analyst:** BKT
ICAL Date: 29-JUN-2016 11:25

			BISMUTH	GERMANIUM	INDIUM
Sample	Type	Run Date	% Rec	% Rec	% Rec
L16060986-03	SAMP	29-JUN-2016 13:13	102.451	99.817	104.214
L16061026-02	SAMP	29-JUN-2016 14:51	89.212	92.118	89.05
L16061026-04	SAMP	29-JUN-2016 14:54	99.491	100.426	97.933
L16061026-06	SAMP	29-JUN-2016 14:57	98.597	101.416	102.107
L16061026-07	SAMP	29-JUN-2016 15:00	108.185	110.897	107.172
WG573548-02	BLANK	29-JUN-2016 12:53	99.243	102.336	95.817
WG573548-03	LCS	29-JUN-2016 12:57	101.717	99.767	101.133
WG574572-01	PSPK	29-JUN-2016 13:16	103.227	99.758	103.895
WG574572-02	SERIAL	29-JUN-2016 13:19	110.008	107.762	106.72
WG574676-05	ICV	29-JUN-2016 11:38	98.257	100.661	96.465
WG574676-06	ICB	29-JUN-2016 11:41	98.914	99.091	98.14
WG574676-07	LLICV	29-JUN-2016 11:46	95.79	95.394	93.877
WG574676-08	ICS	29-JUN-2016 11:50	98.596	98.274	96.165
WG574676-09	ICS	29-JUN-2016 11:54	97.891	98.362	95.991
WG574676-10	CCV	29-JUN-2016 12:00	96.559	97.448	96.281
WG574676-11	CCB	29-JUN-2016 12:07	98.969	97.431	96.855
WG574676-12	CCV	29-JUN-2016 12:37	99.303	102.045	96.376
WG574676-13	CCB	29-JUN-2016 12:40	100.991	102.01	98.548
WG574676-14	LLCCV	29-JUN-2016 12:44	96.535	95.69	95.451
WG574676-15	CCV	29-JUN-2016 13:25	106.544	103.387	103.323
WG574676-16	CCB	29-JUN-2016 13:29	108.277	104.697	104.925
WG574676-17	CCV	29-JUN-2016 14:04	103.528	99.024	101.746
WG574676-18	CCB	29-JUN-2016 14:07	101.206	96.624	98.764
WG574676-19	CCV	29-JUN-2016 15:04	105.397	108.57	106.703
WG574676-20	CCB	29-JUN-2016 15:07	110.064	109.03	109.199
WG574676-27	CCV	29-JUN-2016 17:34	108.39	111.131	108.951
WG574676-28	CCB	29-JUN-2016 17:37	111.551	108.396	114.225
WG574676-29	ICS	29-JUN-2016 17:50	108.567	109.96	110.029
WG574676-30	ICS	29-JUN-2016 17:53	107.111	107.985	110.615
WG574676-31	CCV	29-JUN-2016 17:56	107.683	109.755	111.533
WG574676-32	CCB	29-JUN-2016 17:59	109.386	111.783	111.184
WG574676-33	LLCCV	29-JUN-2016 18:03	109.42	106.253	110.95

Acceptance criteria: 30% - 120% Underlined recoveries are out of range
Acceptance criteria for CCVs and CCBs for method SW846-6020: 80% - 120%

INT_STD_ICPMS - Modified 07/28/2010
PDF File ID: 4832322
Report generated: 07/07/2016 08:33

Microbac Laboratories Inc.

INTERNAL STANDARD REPORT

Login: L16061026 **Analytical Method:** 6020
Analytical Workgroup: WG574572 **Matrix:** 1
Instrument: ICP-MS2 **Analyst:** JYH
ICAL Date: 05-JUL-2016 09:27

			BISMUTH	GERMANIUM	INDIUM
Sample	Type	Run Date	% Rec	% Rec	% Rec
L16061026-02	SAMP	05-JUL-2016 16:05	94.287	91.198	89.78
L16061026-02	SAMP	05-JUL-2016 16:08	97.897	93.356	92.167
L16061026-04	SAMP	05-JUL-2016 16:11	91.832	92.14	89.87
L16061026-04	SAMP	05-JUL-2016 16:15	97.428	92.369	91.144
L16061026-06	SAMP	05-JUL-2016 16:30	95.394	97.869	94.043
L16061026-07	SAMP	05-JUL-2016 16:34	102.173	100.031	99.495
WG573548-02	BLANK	05-JUL-2016 15:49	103.53	101.565	100.369
WG573548-03	LCS	05-JUL-2016 15:52	105.736	104.394	102.796
WG574572-03	PSPK	05-JUL-2016 16:37	102.672	100.652	98.295
WG574572-04	SERIAL	05-JUL-2016 16:41	92.841	89.382	87.188
WG575179-08	ICV	05-JUL-2016 09:50	102.945	104.163	101.103
WG575179-09	ICB	05-JUL-2016 09:53	106.273	107.834	104.639
WG575179-10	LLICV	05-JUL-2016 09:56	103.717	103.348	100.336
WG575179-11	ICS	05-JUL-2016 09:59	100.676	104.095	100.563
WG575179-12	ICS	05-JUL-2016 10:03	104.334	104.793	101.346
WG575179-13	CCV	05-JUL-2016 10:06	100.487	101.459	98.934
WG575179-14	CCB	05-JUL-2016 10:09	106.259	106.237	103.837
WG575179-31	LLCCV	05-JUL-2016 14:36	104.64	97.001	97.387
WG575179-34	CCV	05-JUL-2016 15:41	102.603	101.933	100.295
WG575179-35	CCB	05-JUL-2016 15:44	103.599	100.34	100.262
WG575179-36	CCV	05-JUL-2016 16:18	102.519	101.09	100.867
WG575179-37	CCB	05-JUL-2016 16:21	103.756	100.989	100.122
WG575179-38	CCV	05-JUL-2016 16:47	99.118	97.387	96.41
WG575179-39	CCB	05-JUL-2016 16:50	103.672	101.386	99.679
WG575179-44	CCV	05-JUL-2016 18:36	93.79	92.321	90.717
WG575179-45	CCB	05-JUL-2016 18:39	95.372	91.574	91.3
WG575179-46	ICS	05-JUL-2016 19:08	96.603	92.626	90.775
WG575179-47	ICS	05-JUL-2016 19:11	100.276	99.569	98.291
WG575179-48	CCV	05-JUL-2016 19:14	98.609	96.908	96.301
WG575179-49	CCB	05-JUL-2016 19:17	99.252	96.583	95.361
WG575179-50	LLCCV	05-JUL-2016 19:20	97.791	94.181	93.271

Acceptance criteria: 30% - 120% Underlined recoveries are out of range
Acceptance criteria for CCVs and CCBs for method SW846-6020: 80% - 120%

INT_STD_ICPMS - Modified 07/28/2010
PDF File ID: 4832322
Report generated: 07/07/2016 08:33

Microbac Laboratories Inc.

INTERNAL STANDARD REPORT

Login: L16061026 **Analytical Method:** 6020
Analytical Workgroup: WG574572 **Matrix:** 1
Instrument: ICP-MS2 **Analyst:** JYH
ICAL Date: 06-JUL-2016 09:08

			BISMUTH	GERMANIUM	INDIUM
Sample	Type	Run Date	% Rec	% Rec	% Rec
L16061026-02	SAMP	06-JUL-2016 15:27	98.394	99.815	95.637
WG575370-05	ICV	06-JUL-2016 09:20	100.585	102.463	96.381
WG575370-06	ICB	06-JUL-2016 09:24	100.788	98.692	102.109
WG575370-07	LLICV	06-JUL-2016 09:27	102.41	100.394	95.13
WG575370-08	ICS	06-JUL-2016 09:30	96.414	92.465	92.27
WG575370-09	ICS	06-JUL-2016 09:33	98.737	100.154	96.13
WG575370-10	CCV	06-JUL-2016 09:36	99.702	100.07	99.191
WG575370-11	CCB	06-JUL-2016 09:40	101.529	104.208	96.95
WG575370-26	CCV	06-JUL-2016 14:52	100.655	107.561	102.626
WG575370-27	CCB	06-JUL-2016 14:55	101.918	106.69	102.86
WG575370-28	CCV	06-JUL-2016 15:31	102.288	107.867	104.707
WG575370-29	CCB	06-JUL-2016 15:34	102.911	105.994	103.202
WG575370-38	CCV	06-JUL-2016 18:49	99.508	108.232	103.152
WG575370-39	CCB	06-JUL-2016 18:53	99.139	106.573	101.58
WG575370-40	ICS	06-JUL-2016 19:05	91.725	95.835	90.284
WG575370-41	ICS	06-JUL-2016 19:09	97.884	104.251	99.368
WG575370-42	CCV	06-JUL-2016 19:12	98.906	104.728	100.841
WG575370-43	CCB	06-JUL-2016 19:15	98.402	104.006	99.572
WG575370-44	LLCCV	06-JUL-2016 19:18	98.022	101.982	97.937

Acceptance criteria: 30% - 120% Underlined recoveries are out of range
Acceptance criteria for CCVs and CCBs for method SW846-6020: 80% - 120%

INT_STD_ICPMS - Modified 07/28/2010
PDF File ID: 4832322
Report generated: 07/07/2016 08:33

Microbac Laboratories Inc.
LINEAR RANGE (QUARTERLY)

Login Number: L16061026 **Date:** 04/19/2016
Insturment ID: ICP-MS2 **Method:** 6020A

Analyte	Integration Time (Sec.)	Concentration (ug/L)
Antimony	1.00	100.0
Arsenic	1.00	100.0
Barium	1.00	100.0
Cadmium	1.00	100.0
Chromium	1.00	100.0
Cobalt	1.00	100.0
Copper	1.00	100.0
Lead	1.00	100.0
Manganese	1.00	100.0
Nickel	1.00	100.0
Selenium	1.00	100.0
Silver	1.00	100.0
Thallium	1.00	100.0
Uranium	1.00	100.0
Vanadium	1.00	100.0
Zinc	1.00	100.0

Comments:
All analytes passed acceptance criteria at the specified concentration.

LINEAR_RANGE - Modified 03/06/2008
PDF File ID: 4832321
Report generated: 07/07/2016 08:33

2.3.2.3 Raw Data

MassCal File Name

Mass Calibration File Name Default.tun
MassCal File Path C:\NexIONData\MassCal\Default.tun
Peak Search Window: 1.00

Sample Information

Sample Date/Time: Wednesday, June 29, 2016 09:50:03

Mass Calibration and Resolution

Analyte	E Mass	Meas Mass	Mass C DAC Val	Res DAC Value	Meas Peak W	Custom Res
Li	7.016	6.975	1323	2024	0.699	
Mg	23.985	23.975	4506	2020	0.705	
Co	58.933	58.975	11694	2022	0.697	
In	114.904	114.925	22865	2028	0.690	
U	238.050	238.025	47450	2043	0.688	

Relative Std. Dev.

Mass	Meas. Intens. RSD
5.525	9.254
5.575	1.459
5.625	2.417
5.675	3.442
5.725	2.231
5.775	2.296
5.825	3.784
5.875	3.197
5.925	2.400
5.975	1.087
6.025	1.670
6.075	3.521
6.125	2.277
6.175	2.559
6.225	20.588
6.275	74.689
6.325	108.327
6.375	83.853
6.425	69.722
6.475	32.808
6.525	10.642
6.575	8.985
6.625	1.429
6.675	0.941
6.725	0.857
6.775	3.333
6.825	0.822

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 1

Approved: July 04, 2016

6.875	1.277
6.925	1.899
6.975	1.561
7.025	1.195
7.075	1.059
7.125	1.545
7.175	1.743
7.225	3.248
7.275	8.142
7.325	57.701
7.375	91.287
7.425	51.602
7.475	55.902
7.525	100.000
7.575	69.869
7.625	81.441
7.675	81.441
7.725	70.711
7.775	55.902
7.825	59.266
7.875	49.793
7.925	104.583
7.975	98.543
8.025	94.786
8.075	223.607
8.125	81.441
8.175	84.265
8.225	61.237
8.275	81.312
8.325	81.441
8.375	91.287
8.425	82.402
8.475	22.822
22.525	149.071
22.575	52.705
22.625	46.771
22.675	48.627
22.725	39.087
22.775	18.109
22.825	22.048
22.875	36.422
22.925	32.723
22.975	82.685
23.025	39.239
23.075	25.529
23.125	43.064
23.175	21.317

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 2

Approved: July 04, 2016

23.225	22.232
23.275	15.408
23.325	42.127
23.375	37.172
23.425	24.296
23.475	14.668
23.525	3.807
23.575	1.593
23.625	2.258
23.675	1.180
23.725	1.542
23.775	1.943
23.825	1.185
23.875	1.595
23.925	1.324
23.975	1.055
24.025	1.809
24.075	1.103
24.125	1.249
24.175	0.652
24.225	1.203
24.275	3.614
24.325	39.316
24.375	52.355
24.425	20.738
24.475	5.001
24.525	5.067
24.575	0.930
24.625	1.241
24.675	0.972
24.725	1.552
24.775	0.581
24.825	0.736
24.875	1.684
24.925	0.952
24.975	1.395
25.025	0.762
25.075	1.498
25.125	1.418
25.175	1.145
25.225	6.288
25.275	44.605
25.325	50.990
25.375	73.823
25.425	57.334
25.475	7.010
57.525	8.193

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 3

Approved: July 04, 2016

57.575	3.609
57.625	2.367
57.675	3.747
57.725	3.869
57.775	3.226
57.825	1.406
57.875	2.220
57.925	2.049
57.975	2.397
58.025	3.045
58.075	0.897
58.125	1.397
58.175	1.608
58.225	1.481
58.275	5.244
58.325	6.996
58.375	14.816
58.425	39.994
58.475	18.390
58.525	7.862
58.575	1.876
58.625	1.684
58.675	0.757
58.725	2.431
58.775	2.139
58.825	2.293
58.875	1.987
58.925	2.134
58.975	1.760
59.025	2.893
59.075	1.855
59.125	4.126
59.175	3.545
59.225	2.980
59.275	6.918
59.325	57.198
59.375	79.057
59.425	39.123
59.475	32.983
59.525	7.337
59.575	10.704
59.625	4.888
59.675	5.230
59.725	3.575
59.775	2.279
59.825	3.495
59.875	0.859

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 4

Approved: July 04, 2016

59.925	2.164
59.975	4.187
60.025	3.749
60.075	1.767
60.125	4.373
60.175	1.581
60.225	1.647
60.275	29.768
60.325	42.377
60.375	71.261
60.425	71.261
60.475	69.722
113.525	10.800
113.575	6.260
113.625	1.672
113.675	1.646
113.725	1.420
113.775	2.281
113.825	1.446
113.875	0.820
113.925	1.607
113.975	1.407
114.025	1.046
114.075	2.403
114.125	2.090
114.175	3.331
114.225	5.316
114.275	12.354
114.325	37.872
114.375	26.082
114.425	15.800
114.475	3.831
114.525	4.062
114.575	1.808
114.625	2.980
114.675	3.272
114.725	2.895
114.775	1.904
114.825	1.941
114.875	1.809
114.925	0.906
114.975	0.414
115.025	2.617
115.075	0.808
115.125	1.618
115.175	2.428
115.225	3.339

Report Date/Time: Wednesday, June 29, 2016 09:53:14
Page 5

Approved: July 04, 2016

115.275	5.459
115.325	14.203
115.375	69.869
115.425	66.254
115.475	26.966
115.525	18.235
115.575	7.589
115.625	3.788
115.675	2.661
115.725	3.661
115.775	2.867
115.825	3.952
115.875	6.794
115.925	4.558
115.975	5.421
116.025	1.319
116.075	4.928
116.125	7.259
116.175	4.689
116.225	9.289
116.275	22.491
116.325	26.307
116.375	26.146
116.425	67.711
116.475	40.825
236.525	
236.575	13.258
236.625	39.219
236.675	66.333
236.725	37.815
236.775	20.039
236.825	39.510
236.875	22.934
236.925	27.407
236.975	8.537
237.025	28.853
237.075	45.928
237.125	41.937
237.175	36.377
237.225	26.058
237.275	27.285
237.325	43.196
237.375	45.018
237.425	23.111
237.475	20.908
237.525	26.413
237.575	10.671

Approved: July 04, 2016

237.625	6.558
237.675	4.584
237.725	2.707
237.775	1.019
237.825	0.867
237.875	0.837
237.925	1.315
237.975	0.556
238.025	1.099
238.075	0.473
238.125	0.980
238.175	0.943
238.225	1.564
238.275	1.714
238.325	1.785
238.375	2.057
238.425	2.571
238.475	6.163
238.525	10.798
238.575	23.407
238.625	35.493
238.675	33.171
238.725	15.721
238.775	9.421
238.825	23.516
238.875	27.721
238.925	14.417
238.975	30.010
239.025	24.993
239.075	29.933
239.125	17.678
239.175	38.433
239.225	46.981
239.275	15.215
239.325	48.601
239.375	50.659
239.425	23.570
239.475	38.955

Approved: July 04, 2016

SmartTune Wizard - Summary

```
Optimization Summary

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Start Time: 6/29/2016 10:01:57 AM
End Time: 6/29/2016 10:04:08 AM

Daily Performance Check - [Passed]  Optimum value(s): N/A
        Obtained Intensity (Be 9.0122): 4651.78
        Obtained Intensity (Mg 23.985): 162439.88
        Obtained Intensity (In 114.904): 66202.13
        Obtained Intensity (U 238.05): 140216.56
        Obtained Intensity (Bkgd 220): 0.33
        Obtained Formula (CeO 155.9 / Ce 139.905): 0.018 (=3640.50 / 205153.73)
        Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.004 (=909.89 / 205153.73)
```

Report Date/Time: Wednesday, June 29, 2016 10:04:08
Page 1

Approved: July 04, 2016

SmartTune Wizard - Details

```
Optimization Details

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Optimization Status

Start Time: 6/29/2016 10:01:57 AM

Daily Performance Check
        Optimization Settings:
                Method: C:\NexIONData\Method\ESI Daily Performance.mth.
                Intensity Criterion: Be 9.0122 > 2000
                Intensity Criterion: Mg 23.985 > 15000
                Intensity Criterion: In 114.904 > 40000
                Intensity Criterion: U 238.05 > 30000
                Intensity Criterion: Bkgd 220 <= 5
                Formula Criterion: CeO 155.9 / Ce 139.905 <= 0.025
                Formula Criterion: Ce++ 69.9527 / Ce 139.905 <= 0.03

        Optimization Results:
        Initial Try
                Obtained Intensity (Be 9.0122): 4651.78
                Obtained Intensity (Mg 23.985): 162439.88
                Obtained Intensity (In 114.904): 66202.13
                Obtained Intensity (U 238.05): 140216.56
                Obtained Intensity (Bkgd 220): 0.33
                Obtained Formula (CeO 155.9 / Ce 139.905): 0.018 (=3640.50 / 205153.73)
                Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.004 (=909.89 / 205153.73)
[Passed]  Optimum value(s): N/A

End Time: 6/29/2016 10:04:08 AM
```


Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Blank
Sample Date/Time: Wednesday, June 29, 2016 11:22:34
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74698.8	3.1				ug/L		Standard
	Be	9	8.3	124.9				ug/L		Standard
	Al	27	1493.4	9.1				ug/L		Standard
	Sc	45	61029.4	2.9				ug/L		Standard
	Ti	47	35.0	12.5				ug/L		Standard
	V	51	1434.8	4.3				ug/L		Standard
	Cr	52	7634.6	2.5				ug/L		Standard
	Cr	53	363.3	0.8				ug/L		Standard
	Mn	55	747.0	1.6				ug/L		Standard
	Co	59	233.7	7.1				ug/L		Standard
	Ni	60	214.7	3.5				ug/L		Standard
	Cu	65	104.0	12.0				ug/L		Standard
	Zn	66	363.0	3.8				ug/L		Standard
>	**Ge**	72	495142.3	1.3				ug/L		Standard
	As	75	-49.8	7.8				ug/L		Standard
	Se	82	13.7	23.9				ug/L		Standard
	Se-1	77	42.3	9.8				ug/L		Standard
>	**Ga**	71	41.7	25.0				mg/L		Standard
	Rb	85	16.7	91.7				ug/L		Standard
	Y	89	417098.4	0.9				ug/L		Standard
>	**Rh**	103	13.3	86.6				ug/L		Standard
	Mo	98	10.7	10.9				ug/L		Standard
	Ag	107	68.0	6.4				ug/L		Standard
	Cd	111	2.3	24.9				mg/L		Standard
	Cd	114	11.7	106.5				ug/L		Standard
>	**In**	115	587441.1	0.5				ug/L		Standard
	Sn	118	496.7	9.7				ug/L		Standard
	Sb	123	77.7	34.7				ug/L		Standard
	Ba	135	24.3	19.4				ug/L		Standard
	Ce	140	25.0	20.0				ug/L		Standard
>	**Tb**	159	939701.0	1.8				ug/L		Standard
	Ho	165	10.0	50.0				ug/L		Standard
	Tl	203	4.0	25.0				ug/L		Standard
	Tl	205	1.7	173.2				ug/L		Standard
	Pb	206	239.7	3.6				ug/L		Standard
	Pb	207	197.7	3.9				ug/L		Standard
	Pb	208	883.3	3.3				ug/L		Standard
	U	238	3.3	17.3				ug/L		Standard
>	**Bi**	209	523984.5	1.3				ug/L		Standard

Sample ID: Blank
Report Date/Time: Wednesday, June 29, 2016 11:24:51
Page 1

Approved: July 04, 2016

	Na	23	0.0		mg/L	Standard
	Mg	24	36.7	55.1	mg/L	Standard
	K	39	38.3	52.7	mg/L	Standard
	Ca	43	10.0	50.0	mg/L	Standard
	Fe	54	179.1	14.9	mg/L	Standard
	Fe	57	335.0	10.3	mg/L	Standard
>	**Sc-1**	45	61029.4	2.9	mg/L	Standard
	Cl	35	20194.8	0.8	ug/L	Standard
	Kr	83	4.3	70.5	ug/L	Standard
	Br	81	1546.7	2.4	ug/L	Standard
	P	31	15286.0	3.7	ug/L	Standard
	S	34	1571.7	2.7	ug/L	Standard
	Sr	88	73.3	3.9	ug/L	Standard
	C	12	63.3	24.1	mg/L	Standard
	N	14	0.0		mg/L	Standard
	Hg	202	3.3	173.2	mg/L	Standard
	Dy	164	6.0	187.1	mg/L	Standard
	Ho-1	165	10.0	50.0	mg/L	Standard
	Er	166	13.3	86.6	mg/L	Standard
	I	127	1391.7	17.5	mg/L	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Blank
Report Date/Time: Wednesday, June 29, 2016 11:24:51
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Blank
Report Date/Time: Wednesday, June 29, 2016 11:24:51
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 1
Sample Date/Time: Wednesday, June 29, 2016 11:25:45
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75658.5	2.0				ug/L	74699	Standard
	Be	9	1.7	173.2				ug/L	8	Standard
	Al	27	1555.1	3.1				ug/L	1493	Standard
	Sc	45	60145.9	2.9				ug/L	61029	Standard
	Ti	47	32.7	4.7				ug/L	35	Standard
	V	51	1434.9	4.0				ug/L	1435	Standard
	Cr	52	7544.5	2.5				ug/L	7635	Standard
	Cr	53	365.0	18.0				ug/L	363	Standard
	Mn	55	783.7	0.8				ug/L	747	Standard
	Co	59	239.0	2.9				ug/L	234	Standard
	Ni	60	220.7	1.8				ug/L	215	Standard
	Cu	65	103.0	8.3				ug/L	104	Standard
	Zn	66	365.0	4.1				ug/L	363	Standard
>	**Ge**	72	495351.7	2.1				ug/L	495142	Standard
	As	75	-40.0	8.7				ug/L	-50	Standard
	Se	82	19.1	38.4				ug/L	14	Standard
	Se-1	77	47.3	15.6				ug/L	42	Standard
>	**Ga**	71	53.3	47.2				mg/L	42	Standard
	Rb	85	15.0	88.2				ug/L	17	Standard
	Y	89	427717.4	1.6				ug/L	417098	Standard
>	**Rh**	103	8.3	34.6				ug/L	13	Standard
	Mo	98	11.1	31.8				ug/L	11	Standard
	Ag	107	86.3	19.8				ug/L	68	Standard
	Cd	111	1.3	43.8				mg/L	2	Standard
	Cd	114	3.6	152.7				ug/L	12	Standard
>	**In**	115	599155.0	1.0				ug/L	587441	Standard
	Sn	118	486.7	3.3				ug/L	497	Standard
	Sb	123	28.2	25.8				ug/L	78	Standard
	Ba	135	25.3	13.9				ug/L	24	Standard
	Ce	140	31.7	59.8				ug/L	25	Standard
>	**Tb**	159	985563.2	0.5				ug/L	939701	Standard
	Ho	165	6.7	86.6				ug/L	10	Standard
	Tl	203	1.7	91.7				ug/L	4	Standard
	Tl	205	0.0					ug/L	2	Standard
	Pb	206	249.7	4.0				ug/L	240	Standard
	Pb	207	216.0	2.0				ug/L	198	Standard
	Pb	208	873.0	4.1				ug/L	883	Standard
	U	238	3.3	34.6				ug/L	3	Standard
>	**Bi**	209	535312.9	0.6				ug/L	523984	Standard

Sample ID: Standard 1
Report Date/Time: Wednesday, June 29, 2016 11:28:02
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	mg/L	0	Standard
Mg	24	31.7	18.2	mg/L	37	Standard
K	39	26.7	28.6	mg/L	38	Standard
Ca	43	35.0	24.7	mg/L	10	Standard
Fe	54	130.8	15.0	mg/L	179	Standard
Fe	57	305.0	2.8	mg/L	335	Standard
Sc-1	45	60145.9	2.9	mg/L	61029	Standard
Cl	35	20341.6	1.8	ug/L	20195	Standard
Kr	83	3.0	33.3	ug/L	4	Standard
Br	81	1546.7	10.4	ug/L	1547	Standard
P	31	15145.9	1.8	ug/L	15286	Standard
S	34	1610.1	6.1	ug/L	1572	Standard
Sr	88	86.7	14.5	ug/L	73	Standard
C	12	60.0	28.9	mg/L	63	Standard
N	14	0.0		mg/L	0	Standard
Hg	202	0.0		mg/L	3	Standard
Dy	164	5.7	108.5	mg/L	6	Standard
Ho-1	165	6.7	86.6	mg/L	10	Standard
Er	166	20.0	50.0	mg/L	13	Standard
I	127	1368.4	7.8	mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72			
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: Standard 1
Report Date/Time: Wednesday, June 29, 2016 11:28:02
Page 2

Approved: July 04, 2016

	Rb	85
	Y	89
>	Rh	103
	Mo	98
	Ag	107
	Cd	111
	Cd	114
>	In	115
	Sn	118
	Sb	123
	Ba	135
	Ce	140
>	Tb	159
	Ho	165
	Tl	203
	Tl	205
	Pb	206
	Pb	207
	Pb	208
	U	238
>	Bi	209
	Na	23
	Mg	24
	K	39
	Ca	43
	Fe	54
	Fe	57
>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 1
Report Date/Time: Wednesday, June 29, 2016 11:28:02
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 2
Sample Date/Time: Wednesday, June 29, 2016 11:28:57
Number of Replicates: 3
Autosampler Position: 2
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73081.0	3.5				ug/L	74699	Standard
	Be	9	36.7	56.8				ug/L	8	Standard
	Al	27	9102.7	2.7				ug/L	1493	Standard
	Sc	45	58226.8	3.4				ug/L	61029	Standard
	Ti	47	79.0	2.2				ug/L	35	Standard
	V	51	1848.7	2.3				ug/L	1435	Standard
	Cr	52	7933.1	1.4				ug/L	7635	Standard
	Cr	53	350.0	15.1				ug/L	363	Standard
	Mn	55	1375.4	4.6				ug/L	747	Standard
	Co	59	574.7	5.8				ug/L	234	Standard
	Ni	60	324.3	7.6				ug/L	215	Standard
	Cu	65	285.0	4.9				ug/L	104	Standard
	Zn	66	549.3	0.4				ug/L	363	Standard
>	**Ge**	72	493663.5	1.5				ug/L	495142	Standard
	As	75	-10.4	282.1				ug/L	-50	Standard
	Se	82	20.6	18.5				ug/L	14	Standard
	Se-1	77	49.0	19.5				ug/L	42	Standard
>	**Ga**	71	58.3	27.6				mg/L	42	Standard
	Rb	85	25.0	60.0				ug/L	17	Standard
	Y	89	437555.0	0.8				ug/L	417098	Standard
>	**Rh**	103	25.0	34.6				ug/L	13	Standard
	Mo	98	283.4	4.8				ug/L	11	Standard
	Ag	107	412.3	3.2				ug/L	68	Standard
	Cd	111	121.8	9.1				mg/L	2	Standard
	Cd	114	268.9	5.0				ug/L	12	Standard
>	**In**	115	590951.3	0.9				ug/L	587441	Standard
	Sn	118	850.0	6.4				ug/L	497	Standard
	Sb	123	254.5	2.7				ug/L	78	Standard
	Ba	135	136.0	15.7				ug/L	24	Standard
	Ce	140	18.3	56.8				ug/L	25	Standard
>	**Tb**	159	941723.8	1.9				ug/L	939701	Standard
	Ho	165	3.3	86.6				ug/L	10	Standard
	Tl	203	417.3	5.5				ug/L	4	Standard
	Tl	205	348.3	7.4				ug/L	2	Standard
	Pb	206	633.0	0.8				ug/L	240	Standard
	Pb	207	557.0	2.2				ug/L	198	Standard
	Pb	208	2275.4	1.7				ug/L	883	Standard
	U	238	366.3	0.7				ug/L	3	Standard
>	**Bi**	209	512414.5	1.5				ug/L	523984	Standard

Sample ID: Standard 2
Report Date/Time: Wednesday, June 29, 2016 11:31:14
Page 1

Approved: July 04, 2016

	Na	23	0.0		mg/L	0	Standard
	Mg	24	51.7	24.4	mg/L	37	Standard
	K	39	51.7	24.4	mg/L	38	Standard
	Ca	43	38.3	19.9	mg/L	10	Standard
	Fe	54	160.9	17.5	mg/L	179	Standard
	Fe	57	303.3	9.1	mg/L	335	Standard
>	**Sc-1**	45	58226.8	3.4	mg/L	61029	Standard
	Cl	35	20887.7	1.1	ug/L	20195	Standard
	Kr	83	2.3	107.9	ug/L	4	Standard
	Br	81	1533.4	9.8	ug/L	1547	Standard
	P	31	14952.4	2.3	ug/L	15286	Standard
	S	34	1701.8	4.5	ug/L	1572	Standard
	Sr	88	66.7	31.2	ug/L	73	Standard
	C	12	46.7	65.5	mg/L	63	Standard
	N	14	0.0		mg/L	0	Standard
	Hg	202	0.0		mg/L	3	Standard
	Dy	164	12.7	89.9	mg/L	6	Standard
	Ho-1	165	3.3	86.6	mg/L	10	Standard
	Er	166	13.3	43.3	mg/L	13	Standard
	I	127	1343.4	21.4	mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Standard 2
Report Date/Time: Wednesday, June 29, 2016 11:31:14
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 2
Report Date/Time: Wednesday, June 29, 2016 11:31:14
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 3
Sample Date/Time: Wednesday, June 29, 2016 11:32:08
Number of Replicates: 3
Autosampler Position: 3
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70836.9	0.7				ug/L	74699	Standard
	Be	9	41495.2	4.9	**50.0000**	2.291	4.6	ug/L	8	Standard
	Al	27	6913642.3	1.4	**50.0000**	0.333	0.7	ug/L	1493	Standard
	Sc	45	58109.6	1.6				ug/L	61029	Standard
	Ti	47	38024.7	0.9	**100.0000**	1.896	1.9	ug/L	35	Standard
	V	51	385028.2	1.3	**50.0000**	1.023	2.0	ug/L	1435	Standard
	Cr	52	378864.5	0.8	**50.0000**	0.813	1.6	ug/L	7635	Standard
	Cr	53	46780.4	0.7	**50.0000**	0.853	1.7	ug/L	363	Standard
	Mn	55	354877.0	1.3	**50.0000**	1.137	2.3	ug/L	747	Standard
	Co	59	336932.8	1.2	**50.0000**	1.027	2.1	ug/L	234	Standard
	Ni	60	94633.6	0.8	**50.0000**	0.638	1.3	ug/L	215	Standard
	Cu	65	94252.3	1.0	**50.0000**	0.913	1.8	ug/L	104	Standard
	Zn	66	55983.3	1.5	**50.0000**	1.110	2.2	ug/L	363	Standard
>	**Ge**	72	497274.4	1.0				ug/L	495142	Standard
	As	75	56952.3	0.9	**50.0000**	0.907	1.8	ug/L	-50	Standard
	Se	82	5614.6	2.5	**50.0000**	1.630	3.3	ug/L	14	Standard
	Se-1	77	3760.8	2.5	**50.0000**	1.757	3.5	ug/L	42	Standard
>	**Ga**	71	60.0	22.0				mg/L	42	Standard
	Rb	85	735.0	7.4				ug/L	17	Standard
	Y	89	413691.0	2.8				ug/L	417098	Standard
>	**Rh**	103	28.3	27.0				ug/L	13	Standard
	Mo	98	261684.0	1.3	**100.0000**	1.411	1.4	ug/L	11	Standard
	Ag	107	317580.7	1.1	**50.0000**	0.030	0.1	ug/L	68	Standard
	Cd	111	102024.7	1.3	**50.0000**	0.465	0.9	mg/L	2	Standard
	Cd	114	251691.0	1.5	**50.0000**	1.089	2.2	ug/L	12	Standard
>	**In**	115	567339.3	1.1				ug/L	587441	Standard
	Sn	118	284919.2	1.1	**50.0000**	0.694	1.4	ug/L	497	Standard
	Sb	123	226852.6	0.2	**50.0000**	0.610	1.2	ug/L	78	Standard
	Ba	135	99575.8	0.1	**50.0000**	0.595	1.2	ug/L	24	Standard
	Ce	140	83.3	34.6				ug/L	25	Standard
>	**Tb**	159	932165.5	0.9				ug/L	939701	Standard
	Ho	165	13.3	21.7				ug/L	10	Standard
	Tl	203	405993.6	0.4	**50.0000**	0.749	1.5	ug/L	4	Standard
	Tl	205	349857.0	1.5	**50.0000**	0.282	0.6	ug/L	2	Standard
	Pb	206	265201.9	0.6	**50.0000**	0.817	1.6	ug/L	240	Standard
	Pb	207	235754.7	0.6	**50.0000**	0.829	1.7	ug/L	198	Standard
	Pb	208	926203.5	0.6	**50.0000**	0.626	1.3	ug/L	883	Standard
	U	238	337413.2	1.0	**50.0000**	0.543	1.1	ug/L	3	Standard
>	**Bi**	209	511781.8	1.3				ug/L	523984	Standard

Sample ID: Standard 3
Report Date/Time: Wednesday, June 29, 2016 11:34:25
Page 1

Approved: July 04, 2016

Na	23	0.0					mg/L	0	Standard
Mg	24	3782.1	6.2	**5.0000**	0.388	7.8	mg/L	37	Standard
K	39	2271.8	5.5	**5.0000**	0.254	5.1	mg/L	38	Standard
Ca	43	120.0	23.2	**5.0000**	1.671	33.4	mg/L	10	Standard
Fe	54	5877.0	6.9	**5.0000**	0.418	8.4	mg/L	179	Standard
Fe	57	1703.4	4.4	**5.0000**	0.228	4.6	mg/L	335	Standard
> **Sc-1**	45	58109.6	1.6				mg/L	61029	Standard
Cl	35	22061.4	1.9				ug/L	20195	Standard
Kr	83	2.7	94.4				ug/L	4	Standard
Br	81	1433.4	2.5				ug/L	1547	Standard
P	31	17349.9	2.8				ug/L	15286	Standard
S	34	2026.8	6.8				ug/L	1572	Standard
Sr	88	76.7	26.4				ug/L	73	Standard
C	12	43.3	48.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	33.0	63.6				mg/L	6	Standard
Ho-1	165	13.3	21.7				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1500.1	14.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72			
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: Standard 3
Report Date/Time: Wednesday, June 29, 2016 11:34:25
Page 2

Approved: July 04, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 3
Report Date/Time: Wednesday, June 29, 2016 11:34:25
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: Standard 4
Sample Date/Time: Wednesday, June 29, 2016 11:35:20
Number of Replicates: 3
Autosampler Position: 4
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	70562.3	0.6				ug/L	74699	Standard
	Be	9	83650.3	2.2	**100.5872**	1.689	1.7	ug/L	8	Standard
	Al	27	13996690.0	1.6	**100.8115**	1.548	1.5	ug/L	1493	Standard
	Sc	45	58977.9	1.4				ug/L	61029	Standard
	Ti	47	75710.7	1.4	**200.4046**	3.310	1.7	ug/L	35	Standard
	V	51	745374.0	1.7	**98.8583**	0.465	0.5	ug/L	1435	Standard
	Cr	52	747654.3	1.6	**100.2327**	0.593	0.6	ug/L	7635	Standard
	Cr	53	92243.3	0.8	**99.8509**	0.489	0.5	ug/L	363	Standard
	Mn	55	686210.3	1.7	**98.7774**	1.044	1.1	ug/L	747	Standard
	Co	59	647972.6	0.9	**98.4550**	0.479	0.5	ug/L	234	Standard
	Ni	60	186003.6	1.6	**99.5867**	0.367	0.4	ug/L	215	Standard
	Cu	65	184421.8	1.6	**99.3502**	0.739	0.7	ug/L	104	Standard
	Zn	66	108058.7	1.7	**98.8348**	0.723	0.7	ug/L	363	Standard
>	Ge	72	493312.7	1.3				ug/L	495142	Standard
	As	75	109996.7	2.4	**98.6160**	1.152	1.2	ug/L	-50	Standard
	Se	82	10772.1	0.2	**98.3851**	1.295	1.3	ug/L	14	Standard
	Se-1	77	7226.7	2.3	**98.6814**	1.085	1.1	ug/L	42	Standard
>	Ga	71	88.3	19.9				mg/L	42	Standard
	Rb	85	1528.4	8.2				ug/L	17	Standard
	Y	89	411133.1	2.0				ug/L	417098	Standard
>	Rh	103	50.0	20.0				ug/L	13	Standard
	Mo	98	506691.6	2.0	**196.6866**	6.239	3.2	ug/L	11	Standard
	Ag	107	624459.9	0.9	**99.1146**	1.822	1.8	ug/L	68	Standard
	Cd	111	200021.3	0.6	**98.9639**	1.657	1.7	mg/L	2	Standard
	Cd	114	487602.9	1.3	**98.3617**	2.136	2.2	ug/L	12	Standard
>	In	115	567890.6	1.1				ug/L	587441	Standard
	Sn	118	560558.1	0.7	**99.1782**	1.388	1.4	ug/L	497	Standard
	Sb	123	448476.6	1.3	**99.3779**	2.340	2.4	ug/L	78	Standard
	Ba	135	196432.8	0.6	**99.2744**	1.775	1.8	ug/L	24	Standard
	Ce	140	161.7	46.3				ug/L	25	Standard
>	Tb	159	938936.1	1.2				ug/L	939701	Standard
	Ho	165	10.0					ug/L	10	Standard
	Tl	203	803908.2	0.6	**99.1382**	1.002	1.0	ug/L	4	Standard
	Tl	205	702848.2	0.9	**99.8673**	1.348	1.3	ug/L	2	Standard
	Pb	206	517801.9	1.0	**98.4705**	1.644	1.7	ug/L	240	Standard
	Pb	207	459764.9	1.8	**98.4124**	2.372	2.4	ug/L	198	Standard
	Pb	208	1827348.6	0.4	**98.9940**	0.966	1.0	ug/L	883	Standard
	U	238	674609.9	0.9	**99.6270**	1.321	1.3	ug/L	3	Standard
>	Bi	209	515462.4	0.6				ug/L	523984	Standard

Sample ID: Standard 4
Report Date/Time: Wednesday, June 29, 2016 11:37:37
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.9850**	34.606	173.2	mg/L	0	Standard
Mg	24	7435.2	2.2	**9.8676**	0.257	2.6	mg/L	37	Standard
K	39	4282.3	4.3	**9.6823**	0.481	5.0	mg/L	38	Standard
Ca	43	178.3	18.2	**9.1297**	2.109	23.1	mg/L	10	Standard
Fe	54	11328.6	1.8	**9.8033**	0.307	3.1	mg/L	179	Standard
Fe	57	3145.3	5.1	**9.9880**	0.532	5.3	mg/L	335	Standard
Sc-1	45	58977.9	1.4				mg/L	61029	Standard
Cl	35	21244.2	2.5				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	1583.4	3.9				ug/L	1547	Standard
P	31	17550.2	2.6				ug/L	15286	Standard
S	34	2090.1	3.5				ug/L	1572	Standard
Sr	88	83.3	42.6				ug/L	73	Standard
C	12	56.7	62.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	22.9	66.8				mg/L	6	Standard
Ho-1	165	10.0					mg/L	10	Standard
Er	166	10.0					mg/L	13	Standard
I	127	1638.4	19.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72			
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: Standard 4
Report Date/Time: Wednesday, June 29, 2016 11:37:37
Page 2

Approved: July 04, 2016

	Rb	85
	Y	89
>	Rh	103
	Mo	98
	Ag	107
	Cd	111
	Cd	114
>	In	115
	Sn	118
	Sb	123
	Ba	135
	Ce	140
>	Tb	159
	Ho	165
	Tl	203
	Tl	205
	Pb	206
	Pb	207
	Pb	208
	U	238
>	Bi	209
	Na	23
	Mg	24
	K	39
	Ca	43
	Fe	54
	Fe	57
>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Corr. Coef.	Na	23	Correlation coefficient < 0.998
Corr. Coef.	Ca	43	Correlation coefficient < 0.998

Sample ID: Standard 4
Report Date/Time: Wednesday, June 29, 2016 11:37:37
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 1
Sample Date/Time: Wednesday, June 29, 2016 11:38:33
Number of Replicates: 3
Autosampler Position: 201
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	69383.6	2.2				ug/L	74699	Standard
	Be	9	41665.5	1.0	**50.9761**	1.262	2.5	ug/L	8	Standard
	Al	27	7007266.1	0.7	**51.3351**	1.064	2.1	ug/L	1493	Standard
	Sc	45	57040.5	0.5				ug/L	61029	Standard
	Ti	47	39782.2	0.5	**104.1738**	1.427	1.4	ug/L	35	Standard
	V	51	390724.1	1.4	**51.2089**	1.334	2.6	ug/L	1435	Standard
	Cr	52	380689.4	0.9	**50.0149**	1.033	2.1	ug/L	7635	Standard
	Cr	53	46596.6	2.0	**49.7615**	1.193	2.4	ug/L	363	Standard
	Mn	55	364452.5	1.9	**51.8606**	1.232	2.4	ug/L	747	Standard
	Co	59	335357.4	2.0	**50.4145**	0.943	1.9	ug/L	234	Standard
	Ni	60	95528.2	0.6	**50.5685**	0.768	1.5	ug/L	215	Standard
	Cu	65	95355.8	0.8	**50.7995**	0.891	1.8	ug/L	104	Standard
	Zn	66	57795.4	0.9	**52.1095**	0.303	0.6	ug/L	363	Standard
>	Ge	72	498415.5	1.2				ug/L	495142	Standard
	As	75	57290.2	0.4	**50.8755**	0.674	1.3	ug/L	-50	Standard
	Se	82	5486.5	1.1	**49.5278**	0.838	1.7	ug/L	14	Standard
	Se-1	77	3797.8	2.0	**51.0456**	1.569	3.1	ug/L	42	Standard
>	Ga	71	88.3	43.2				mg/L	42	Standard
	Rb	85	576.7	2.8				ug/L	17	Standard
	Y	89	413113.3	1.2				ug/L	417098	Standard
>	Rh	103	26.7	92.5				ug/L	13	Standard
	Mo	98	275236.9	1.8	**107.0386**	1.474	1.4	ug/L	11	Standard
	Ag	107	320241.4	0.9	**50.9276**	0.747	1.5	ug/L	68	Standard
	Cd	111	101502.5	0.4	**50.3185**	0.332	0.7	mg/L	2	Standard
	Cd	114	248344.9	2.5	**50.2016**	1.511	3.0	ug/L	12	Standard
>	In	115	566676.1	0.5				ug/L	587441	Standard
	Sn	118	310270.4	1.9	**54.9671**	1.097	2.0	ug/L	497	Standard
	Sb	123	224427.6	0.4	**49.8273**	0.118	0.2	ug/L	78	Standard
	Ba	135	98607.5	0.6	**49.9289**	0.574	1.1	ug/L	24	Standard
	Ce	140	138.3	11.0				ug/L	25	Standard
>	Tb	159	931841.4	1.3				ug/L	939701	Standard
	Ho	165	10.0	86.6				ug/L	10	Standard
	Tl	203	410286.1	0.9	**50.6578**	0.710	1.4	ug/L	4	Standard
	Tl	205	350970.1	0.3	**49.9294**	0.473	0.9	ug/L	2	Standard
	Pb	206	272426.4	0.6	**51.8371**	0.812	1.6	ug/L	240	Standard
	Pb	207	234103.7	0.5	**50.1331**	0.525	1.0	ug/L	198	Standard
	Pb	208	940824.8	0.3	**50.9948**	0.649	1.3	ug/L	883	Standard
	U	238	333596.8	0.3	**49.3228**	0.442	0.9	ug/L	3	Standard
>	Bi	209	514849.0	1.0				ug/L	523984	Standard

Sample ID: QC Std 1
Report Date/Time: Wednesday, June 29, 2016 11:40:50
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	3822.1	2.3	**5.2142**	0.148	2.8	mg/L	37	Standard
K	39	2126.8	6.1	**4.9163**	0.328	6.7	mg/L	38	Standard
Ca	43	118.3	24.4	**5.4681**	1.919	35.1	mg/L	10	Standard
Fe	54	5948.1	0.7	**5.2575**	0.061	1.2	mg/L	179	Standard
Fe	57	1791.8	1.9	**5.4420**	0.142	2.6	mg/L	335	Standard
> **Sc-1**	45	57040.5	0.5				mg/L	61029	Standard
Cl	35	20895.1	0.7				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	1520.1	6.3				ug/L	1547	Standard
P	31	17665.3	1.6				ug/L	15286	Standard
S	34	2033.5	1.7				ug/L	1572	Standard
Sr	88	80.0	10.8				ug/L	73	Standard
C	12	70.0	24.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	23.0	24.5				mg/L	6	Standard
Ho-1	165	10.0	86.6				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1353.4	28.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	101.952		
Al	27	102.670		
Sc	45			
Ti	47	104.174		
V	51	102.418		
Cr	52	100.030		
Cr	53			
Mn	55	103.721		
Co	59	100.829		
Ni	60	101.137		
Cu	65	101.599		
Zn	66	104.219		
> Ge	72		100.661	
As	75	101.751		
Se	82	99.056		
Se-1	77			
> Ga	71			

Sample ID: QC Std 1
Report Date/Time: Wednesday, June 29, 2016 11:40:50
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	107.039		
Ag	107	101.855		
Cd	111	100.637		
Cd	114			
In	115		96.465	
Sn	118	109.934		
Sb	123	99.655		
Ba	135	99.858		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.316		
Tl	205			
Pb	206	103.674		
Pb	207	100.266		
Pb	208	101.990		
U	238	98.646		
Bi	209		98.257	
Na	23	0.100		
Mg	24	104.283		
K	39	98.326		
Ca	43	109.361		
Fe	54	105.151		
Fe	57	108.840		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 1	Na	23	
QC Std 1	Ca	43	

Sample ID: QC Std 1
Report Date/Time: Wednesday, June 29, 2016 11:40:50
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 2
Sample Date/Time: Wednesday, June 29, 2016 11:41:46
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	70378.1	2.5				ug/L	74699	Standard
	Be	9	20.0	25.0	**0.0316**	0.005	17.2	ug/L	8	Standard
	Al	27	2221.8	11.1	**0.0027**	0.002	62.7	ug/L	1493	Standard
	Sc	45	58138.1	3.2				ug/L	61029	Standard
	Ti	47	47.0	20.3	**0.0161**	0.026	158.4	ug/L	35	Standard
	V	51	1216.9	5.3	**-0.0330**	0.008	24.4	ug/L	1435	Standard
	Cr	52	6478.4	1.5	**-0.1416**	0.009	6.0	ug/L	7635	Standard
	Cr	53	358.3	10.7	**0.0617**	0.046	74.0	ug/L	363	Standard
	Mn	55	785.0	2.1	**-0.0343**	0.003	7.5	ug/L	747	Standard
	Co	59	292.7	8.2	**0.0075**	0.004	50.2	ug/L	234	Standard
	Ni	60	209.7	4.3	**-0.0109**	0.006	51.3	ug/L	215	Standard
	Cu	65	121.0	5.8	**-0.0381**	0.003	8.4	ug/L	104	Standard
	Zn	66	398.3	6.4	**-0.0865**	0.023	26.3	ug/L	363	Standard
>	**Ge**	72	490643.1	1.0				ug/L	495142	Standard
	As	75	-2.3	851.2	**0.0577**	0.017	30.2	ug/L	-50	Standard
	Se	82	15.9	39.8	**0.0087**	0.059	680.3	ug/L	14	Standard
	Se-1	77	51.3	11.1	**0.0869**	0.077	88.1	ug/L	42	Standard
>	**Ga**	71	46.7	16.4				mg/L	42	Standard
	Rb	85	33.3	45.8				ug/L	17	Standard
	Y	89	424498.7	1.6				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	405.0	23.6	**0.1487**	0.033	22.3	ug/L	11	Standard
	Ag	107	128.3	5.2	**0.0071**	0.001	8.6	ug/L	68	Standard
	Cd	111	15.0	38.7	**-0.0006**	0.003	474.1	mg/L	2	Standard
	Cd	114	47.2	8.1	**0.0073**	0.001	11.6	ug/L	12	Standard
>	**In**	115	576512.9	2.1				ug/L	587441	Standard
	Sn	118	851.7	19.9	**0.0543**	0.032	58.8	ug/L	497	Standard
	Sb	123	1550.0	28.1	**0.3352**	0.100	29.7	ug/L	78	Standard
	Ba	135	36.7	23.2	**0.0022**	0.004	189.8	ug/L	24	Standard
	Ce	140	21.7	93.3				ug/L	25	Standard
>	**Tb**	159	924023.3	3.2				ug/L	939701	Standard
	Ho	165	10.0	100.0				ug/L	10	Standard
	Tl	203	60.7	18.1	**0.0057**	0.001	22.9	ug/L	4	Standard
	Tl	205	61.7	61.4	**0.0089**	0.005	59.3	ug/L	2	Standard
	Pb	206	271.3	7.0	**-0.0198**	0.004	19.8	ug/L	240	Standard
	Pb	207	237.3	7.1	**-0.0194**	0.004	21.0	ug/L	198	Standard
	Pb	208	1028.3	6.0	**-0.0187**	0.003	15.4	ug/L	883	Standard
	U	238	47.7	28.5	**0.0026**	0.002	78.0	ug/L	3	Standard
>	**Bi**	209	518292.4	0.9				ug/L	523984	Standard

Sample ID: QC Std 2
Report Date/Time: Wednesday, June 29, 2016 11:44:03
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	35.0	42.9	**-0.0179**	0.019	107.8	mg/L	37	Standard
K	39	36.7	39.4	**-0.0284**	0.035	124.9	mg/L	38	Standard
Ca	43	25.0	40.0	**-0.8569**	0.683	79.6	mg/L	10	Standard
Fe	54	125.6	8.4	**-0.0262**	0.013	49.1	mg/L	179	Standard
Fe	57	296.7	4.9	**-0.0150**	0.061	408.2	mg/L	335	Standard
> **Sc-1**	45	58138.1	3.2				mg/L	61029	Standard
Cl	35	20966.5	2.9				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	1540.1	1.7				ug/L	1547	Standard
P	31	17027.9	2.4				ug/L	15286	Standard
S	34	1976.8	0.6				ug/L	1572	Standard
Sr	88	85.0	5.9				ug/L	73	Standard
C	12	30.0	33.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	6.3	86.9				mg/L	6	Standard
Ho-1	165	10.0	100.0				mg/L	10	Standard
Er	166	6.7	173.2				mg/L	13	Standard
I	127	1458.4	25.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.091	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 2
Report Date/Time: Wednesday, June 29, 2016 11:44:03
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.140
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	98.914
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 2	Sb	123	
QC Std 2	Ca	43	

Sample ID: QC Std 2
Report Date/Time: Wednesday, June 29, 2016 11:44:03
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 3
Sample Date/Time: Wednesday, June 29, 2016 11:46:53
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	66549.2	2.6				ug/L	74699	Standard
	Be	9	181.7	3.2	**0.2393**	0.009	3.6	ug/L	8	Standard
	Al	27	1430.1	59.0	**-0.0024**	0.007	267.2	ug/L	1493	Standard
	Sc	45	54693.6	2.6				ug/L	61029	Standard
	Ti	47	40.3	11.5	**0.0025**	0.013	506.9	ug/L	35	Standard
	V	51	4161.2	3.6	**0.3819**	0.016	4.1	ug/L	1435	Standard
	Cr	52	12463.1	1.9	**0.7390**	0.018	2.5	ug/L	7635	Standard
	Cr	53	996.7	1.0	**0.8006**	0.013	1.6	ug/L	363	Standard
	Mn	55	4091.6	2.3	**0.4679**	0.008	1.8	ug/L	747	Standard
	Co	59	2840.6	1.4	**0.4137**	0.010	2.4	ug/L	234	Standard
	Ni	60	3014.0	1.2	**1.5638**	0.035	2.2	ug/L	215	Standard
	Cu	65	1546.7	1.4	**0.7676**	0.020	2.7	ug/L	104	Standard
	Zn	66	7394.5	0.3	**6.6420**	0.044	0.7	ug/L	363	Standard
>	**Ge**	72	472336.8	0.9				ug/L	495142	Standard
	As	75	394.7	6.3	**0.4290**	0.024	5.7	ug/L	-50	Standard
	Se	82	54.4	27.0	**0.3818**	0.143	37.5	ug/L	14	Standard
	Se-1	77	75.3	14.2	**0.4599**	0.162	35.3	ug/L	42	Standard
>	**Ga**	71	58.3	44.0				mg/L	42	Standard
	Rb	85	8.3	34.6				ug/L	17	Standard
	Y	89	402582.8	1.9				ug/L	417098	Standard
>	**Rh**	103	5.0	100.0				ug/L	13	Standard
	Mo	98	114.4	34.0	**0.0400**	0.016	39.3	ug/L	11	Standard
	Ag	107	2464.9	1.7	**0.3900**	0.008	2.1	ug/L	68	Standard
	Cd	111	479.8	5.6	**0.2365**	0.013	5.7	mg/L	2	Standard
	Cd	114	1217.0	2.5	**0.2507**	0.007	2.6	ug/L	12	Standard
>	**In**	115	551471.6	0.4				ug/L	587441	Standard
	Sn	118	548.3	23.3	**0.0053**	0.024	443.1	ug/L	497	Standard
	Sb	123	2017.6	6.1	**0.4562**	0.029	6.4	ug/L	78	Standard
	Ba	135	1439.7	3.7	**0.7333**	0.029	4.0	ug/L	24	Standard
	Ce	140	11.7	49.5				ug/L	25	Standard
>	**Tb**	159	887425.9	1.8				ug/L	939701	Standard
	Ho	165	10.0	50.0				ug/L	10	Standard
	Tl	203	701.7	6.6	**0.0871**	0.005	5.3	ug/L	4	Standard
	Tl	205	588.3	14.1	**0.0861**	0.013	15.2	ug/L	2	Standard
	Pb	206	1333.4	2.9	**0.1894**	0.006	3.1	ug/L	240	Standard
	Pb	207	1083.0	1.4	**0.1683**	0.005	3.1	ug/L	198	Standard
	Pb	208	4575.3	2.2	**0.1806**	0.004	2.0	ug/L	883	Standard
	U	238	2677.9	3.2	**0.4017**	0.012	2.9	ug/L	3	Standard
>	**Bi**	209	501924.0	1.5				ug/L	523984	Standard

Sample ID: QC Std 3
Report Date/Time: Wednesday, June 29, 2016 11:49:09
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**21.9523**	38.014	173.2	mg/L	0	Standard
Mg	24	43.3	29.0	**-0.0023**	0.020	841.8	mg/L	37	Standard
K	39	33.3	37.7	**-0.0317**	0.032	101.1	mg/L	38	Standard
Ca	43	46.7	12.4	**0.7738**	0.482	62.2	mg/L	10	Standard
Fe	54	100.7	18.6	**-0.0432**	0.016	36.6	mg/L	179	Standard
Fe	57	325.0	8.1	**0.1578**	0.085	53.5	mg/L	335	Standard
> **Sc-1**	45	54693.6	2.6				mg/L	61029	Standard
Cl	35	21199.5	1.6				ug/L	20195	Standard
Kr	83	3.3	17.3				ug/L	4	Standard
Br	81	1403.4	5.3				ug/L	1547	Standard
P	31	11833.0	1.6				ug/L	15286	Standard
S	34	2000.1	8.0				ug/L	1572	Standard
Sr	88	93.3	20.3				ug/L	73	Standard
C	12	56.7	10.2				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	13.0	89.8				mg/L	6	Standard
Ho-1	165	10.0	50.0				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1430.1	34.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	119.656		
Al	27	-0.244		
Sc	45			
Ti	47			
V	51	95.473		
Cr	52	92.381		
Cr	53			
Mn	55	93.583		
Co	59	103.429		
Ni	60	97.737		
Cu	65	95.951		
Zn	66	106.273		
> Ge	72		95.394	
As	75	107.259		
Se	82	95.456		
Se-1	77			
> Ga	71			

Sample ID: QC Std 3
Report Date/Time: Wednesday, June 29, 2016 11:49:09
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	97.496		
Cd	111	98.556		
Cd	114			
> In	115			93.877
Sn	118			
Sb	123	114.045		
Ba	135	97.779		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	108.827		
Tl	205			
Pb	206			
Pb	207			
Pb	208	90.316		
U	238	100.429		
> Bi	209			95.790
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 3	Al	27	

Sample ID: QC Std 3
Report Date/Time: Wednesday, June 29, 2016 11:49:09
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Wednesday, June 29, 2016 11:50:53
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	71628.9	1.8				ug/L	74699	Standard
	Be	9	13.3	43.3	**0.0235**	0.007	30.3	ug/L	8	Standard
	Al	27	7079214.4	1.4	**50.2405**	1.571	3.1	ug/L	1493	Standard
	Sc	45	57092.4	1.6				ug/L	61029	Standard
	Ti	47	41841.7	1.2	**112.2257**	1.296	1.2	ug/L	35	Standard
	V	51	1975.7	11.7	**0.0706**	0.031	44.3	ug/L	1435	Standard
	Cr	52	7951.4	1.4	**0.0680**	0.015	21.7	ug/L	7635	Standard
	Cr	53	3292.0	3.5	**3.2946**	0.126	3.8	ug/L	363	Standard
	Mn	55	1443.4	4.1	**0.0629**	0.009	13.6	ug/L	747	Standard
	Co	59	431.3	5.1	**0.0292**	0.003	11.6	ug/L	234	Standard
	Ni	60	837.4	1.6	**0.3312**	0.007	2.2	ug/L	215	Standard
	Cu	65	638.0	4.2	**0.2451**	0.015	6.0	ug/L	104	Standard
	Zn	66	1716.8	2.4	**1.1448**	0.039	3.4	ug/L	363	Standard
>	Ge	72	486594.7	0.0				ug/L	495142	Standard
	As	75	-47.8	111.3	**0.0161**	0.048	300.1	ug/L	-50	Standard
	Se	82	12.5	17.3	**-0.0221**	0.020	90.2	ug/L	14	Standard
	Se-1	77	285.3	3.8	**3.3530**	0.150	4.5	ug/L	42	Standard
>	Ga	71	146.7	5.2				mg/L	42	Standard
	Rb	85	1056.7	1.8				ug/L	17	Standard
	Y	89	409909.2	0.2				ug/L	417098	Standard
>	Rh	103	11.7	99.0				ug/L	13	Standard
	Mo	98	258876.2	1.4	**101.0110**	2.466	2.4	ug/L	11	Standard
	Ag	107	250.7	8.2	**0.0271**	0.004	13.5	ug/L	68	Standard
	Cd	111	-191.4	6.3	**-0.1031**	0.005	5.1	mg/L	2	Standard
	Cd	114	476.2	9.0	**0.0945**	0.009	9.4	ug/L	12	Standard
>	In	115	564912.9	1.1				ug/L	587441	Standard
	Sn	118	765.0	2.0	**0.0415**	0.002	5.8	ug/L	497	Standard
	Sb	123	314.8	31.7	**0.0658**	0.022	32.8	ug/L	78	Standard
	Ba	135	628.0	0.0	**0.3031**	0.004	1.2	ug/L	24	Standard
	Ce	140	243.3	10.5				ug/L	25	Standard
>	Tb	159	934364.8	1.3				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	234.7	3.6	**0.0271**	0.001	3.8	ug/L	4	Standard
	Tl	205	175.0	22.3	**0.0250**	0.006	22.8	ug/L	2	Standard
	Pb	206	514.7	3.8	**0.0265**	0.004	16.6	ug/L	240	Standard
	Pb	207	427.7	5.6	**0.0214**	0.005	22.5	ug/L	198	Standard
	Pb	208	1691.4	2.8	**0.0174**	0.002	11.4	ug/L	883	Standard
	U	238	20.7	10.1	**-0.0014**	0.000	22.7	ug/L	3	Standard
>	Bi	209	516629.8	0.9				ug/L	523984	Standard

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 11:53:10
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**102.8074**	36.030	35.0	mg/L	0	Standard
Mg	24	9604.7	1.9	**13.1909**	0.366	2.8	mg/L	37	Standard
K	39	2033.5	6.6	**4.6893**	0.286	6.1	mg/L	38	Standard
Ca	43	250.0	4.0	**14.3615**	0.772	5.4	mg/L	10	Standard
Fe	54	14441.1	1.7	**12.9538**	0.416	3.2	mg/L	179	Standard
Fe	57	3845.5	3.6	**12.9047**	0.714	5.5	mg/L	335	Standard
Sc-1	45	57092.4	1.6				mg/L	61029	Standard
Cl	35	20428.4	2.1				ug/L	20195	Standard
Kr	83	3.0	57.7				ug/L	4	Standard
Br	81	1493.4	8.7				ug/L	1547	Standard
P	31	15903.3	2.0				ug/L	15286	Standard
S	34	1818.4	8.5				ug/L	1572	Standard
Sr	88	70.0	35.7				ug/L	73	Standard
C	12	136.7	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	46.7	86.6				mg/L	3	Standard
Dy	164	12.2	93.4				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	23.3	24.7				mg/L	13	Standard
I	127	2226.8	19.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27	1.005		
Sc	45			
Ti	47	112.226		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		98.274	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 11:53:10
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	101.011		
Ag	107			
Cd	111			
Cd	114			
> In	115		96.165	
Sn	118			
Sb	123			
Ba	135			
Ce	140			
> Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
> Bi	209		98.596	
Na	23	822.459		
Mg	24	263.819		
K	39	93.787		
Ca	43	95.743		
Fe	54	103.630		
Fe	57	103.238		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Na	23	
QC Std 4	Mg	24	

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 11:53:10
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Wednesday, June 29, 2016 11:54:47
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	72764.4	2.4				ug/L	74699	Standard
	Be	9	86093.9	0.6	**100.4334**	2.352	2.3	ug/L	8	Standard
	Al	27	7577089.0	1.5	**52.9410**	1.751	3.3	ug/L	1493	Standard
	Sc	45	54926.1	3.1				ug/L	61029	Standard
	Ti	47	43156.4	1.1	**115.6496**	0.651	0.6	ug/L	35	Standard
	V	51	767037.5	0.7	**103.0563**	0.476	0.5	ug/L	1435	Standard
	Cr	52	743370.8	0.7	**100.9527**	0.091	0.1	ug/L	7635	Standard
	Cr	53	95736.5	0.4	**104.9834**	0.488	0.5	ug/L	363	Standard
	Mn	55	719673.0	1.2	**104.9487**	1.961	1.9	ug/L	747	Standard
	Co	59	684381.8	1.6	**105.3392**	2.464	2.3	ug/L	234	Standard
	Ni	60	184775.9	0.8	**100.2112**	0.878	0.9	ug/L	215	Standard
	Cu	65	183945.6	0.7	**100.3790**	1.285	1.3	ug/L	104	Standard
	Zn	66	111716.4	0.5	**103.5231**	0.465	0.4	ug/L	363	Standard
>	**Ge**	72	487033.1	0.8				ug/L	495142	Standard
	As	75	115123.8	0.6	**104.5507**	0.234	0.2	ug/L	-50	Standard
	Se	82	11162.3	1.3	**103.2681**	1.937	1.9	ug/L	14	Standard
	Se-1	77	7790.3	1.4	**107.8133**	0.940	0.9	ug/L	42	Standard
>	**Ga**	71	155.0	24.4				mg/L	42	Standard
	Rb	85	853.4	12.0				ug/L	17	Standard
	Y	89	414937.1	1.1				ug/L	417098	Standard
>	**Rh**	103	48.3	21.5				ug/L	13	Standard
	Mo	98	271334.5	0.2	**106.0489**	0.711	0.7	ug/L	11	Standard
	Ag	107	596492.9	1.3	**95.3413**	1.752	1.8	ug/L	68	Standard
	Cd	111	202245.3	1.2	**100.7713**	1.934	1.9	mg/L	2	Standard
	Cd	114	489634.1	0.4	**99.4602**	0.649	0.7	ug/L	12	Standard
>	**In**	115	563893.1	0.8				ug/L	587441	Standard
	Sn	118	898.4	12.5	**0.0654**	0.019	28.7	ug/L	497	Standard
	Sb	123	448934.9	1.4	**100.1777**	2.039	2.0	ug/L	78	Standard
	Ba	135	193320.5	0.8	**98.3885**	1.494	1.5	ug/L	24	Standard
	Ce	140	68.3	16.9				ug/L	25	Standard
>	**Tb**	159	922182.5	1.6				ug/L	939701	Standard
	Ho	165	15.0	57.7				ug/L	10	Standard
	Tl	203	808287.0	0.7	**100.1763**	1.478	1.5	ug/L	4	Standard
	Tl	205	710155.2	1.4	**101.4182**	2.560	2.5	ug/L	2	Standard
	Pb	206	542765.1	0.5	**103.7300**	0.986	1.0	ug/L	240	Standard
	Pb	207	466015.8	1.0	**100.2432**	1.559	1.6	ug/L	198	Standard
	Pb	208	1879754.6	0.3	**102.3429**	1.262	1.2	ug/L	883	Standard
	U	238	673426.3	1.1	**99.9548**	2.202	2.2	ug/L	3	Standard
>	**Bi**	209	512933.6	1.2				ug/L	523984	Standard

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 11:57:04
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	10138.4	1.8	**14.4874**	0.596	4.1	mg/L	37	Standard
K	39	2125.1	3.3	**5.1060**	0.147	2.9	mg/L	38	Standard
Ca	43	265.0	8.2	**16.1201**	2.142	13.3	mg/L	10	Standard
Fe	54	14876.0	0.8	**13.8828**	0.353	2.5	mg/L	179	Standard
Fe	57	3958.9	4.0	**13.8986**	1.069	7.7	mg/L	335	Standard
> **Sc-1**	45	54926.1	3.1				mg/L	61029	Standard
Cl	35	20351.0	1.2				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	1600.1	9.2				ug/L	1547	Standard
P	31	14895.7	0.4				ug/L	15286	Standard
S	34	1618.4	8.3				ug/L	1572	Standard
Sr	88	85.0	10.2				ug/L	73	Standard
C	12	116.7	71.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	16.2	38.3				mg/L	6	Standard
Ho-1	165	15.0	57.7				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	1496.7	25.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	100.433		
Al	27	1.059		
Sc	45			
Ti	47	115.650		
V	51	103.056		
Cr	52	100.953		
Cr	53			
Mn	55	104.949		
Co	59	105.339		
Ni	60	100.211		
Cu	65	100.379		
Zn	66	103.523		
> Ge	72		98.362	
As	75	104.551		
Se	82	103.268		
Se-1	77			
> Ga	71			

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 11:57:04
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	106.049		
Ag	107	95.341		
Cd	111	100.771		
Cd	114			
> In	115		95.991	
Sn	118			
Sb	123	100.178		
Ba	135	98.389		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	100.176		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.343		
U	238	99.955		
> Bi	209		97.891	
Na	23	0.040		
Mg	24	289.748		
K	39	102.121		
Ca	43	107.468		
Fe	54	111.063		
Fe	57	111.189		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Na	23	
QC Std 5	Mg	24	

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 11:57:04
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 12:00:10
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	74127.8	4.1				ug/L	74699	Standard
	Be	9	44057.3	2.0	**50.4723**	1.532	3.0	ug/L	8	Standard
	Al	27	7182862.5	0.8	**49.2846**	1.691	3.4	ug/L	1493	Standard
	Sc	45	54972.9	1.9				ug/L	61029	Standard
	Ti	47	37035.5	1.0	**100.2235**	3.375	3.4	ug/L	35	Standard
	V	51	380987.7	1.2	**51.6037**	1.920	3.7	ug/L	1435	Standard
	Cr	52	371095.9	0.5	**50.3905**	1.647	3.3	ug/L	7635	Standard
	Cr	53	45648.7	0.9	**50.3715**	0.927	1.8	ug/L	363	Standard
	Mn	55	362758.6	1.1	**53.3556**	2.033	3.8	ug/L	747	Standard
	Co	59	346170.1	1.6	**53.7955**	2.203	4.1	ug/L	234	Standard
	Ni	60	91822.1	1.5	**50.2407**	2.119	4.2	ug/L	215	Standard
	Cu	65	92043.5	2.0	**50.6873**	2.359	4.7	ug/L	104	Standard
	Zn	66	54708.9	1.1	**50.9762**	1.937	3.8	ug/L	363	Standard
>	Ge	72	482505.9	2.7				ug/L	495142	Standard
	As	75	56019.2	1.2	**51.4117**	1.775	3.5	ug/L	-50	Standard
	Se	82	5615.5	1.4	**52.4035**	2.100	4.0	ug/L	14	Standard
	Se-1	77	3787.5	1.5	**52.6293**	2.112	4.0	ug/L	42	Standard
>	Ga	71	63.3	9.1				mg/L	42	Standard
	Rb	85	771.7	5.9				ug/L	17	Standard
	Y	89	419939.4	2.7				ug/L	417098	Standard
>	Rh	103	31.7	9.1				ug/L	13	Standard
	Mo	98	262674.3	2.0	**102.5093**	6.034	5.9	ug/L	11	Standard
	Ag	107	318557.4	0.8	**50.8058**	2.007	4.0	ug/L	68	Standard
	Cd	111	103808.6	0.6	**51.6156**	2.157	4.2	mg/L	2	Standard
	Cd	114	253134.5	2.1	**51.2889**	1.034	2.0	ug/L	12	Standard
>	In	115	565595.5	4.0				ug/L	587441	Standard
	Sn	118	290084.1	1.0	**51.5381**	2.174	4.2	ug/L	497	Standard
	Sb	123	228198.9	1.3	**50.8307**	2.651	5.2	ug/L	78	Standard
	Ba	135	99366.1	0.6	**50.4579**	1.943	3.9	ug/L	24	Standard
	Ce	140	78.3	13.3				ug/L	25	Standard
>	Tb	159	913248.6	2.6				ug/L	939701	Standard
	Ho	165	3.3	86.6				ug/L	10	Standard
	Tl	203	413066.5	1.0	**51.9084**	1.016	2.0	ug/L	4	Standard
	Tl	205	363184.8	0.9	**52.5982**	1.678	3.2	ug/L	2	Standard
	Pb	206	268800.2	0.8	**52.0649**	1.553	3.0	ug/L	240	Standard
	Pb	207	239468.7	1.2	**52.2107**	1.790	3.4	ug/L	198	Standard
	Pb	208	958802.6	0.9	**52.9068**	1.662	3.1	ug/L	883	Standard
	U	238	349772.5	0.2	**52.6413**	1.317	2.5	ug/L	3	Standard
>	Bi	209	505954.2	2.3				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:02:27
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	3805.5	5.3	**5.3881**	0.260	4.8	mg/L	37	Standard
K	39	2106.8	1.9	**5.0574**	0.179	3.5	mg/L	38	Standard
Ca	43	101.7	5.7	**4.6110**	0.501	10.9	mg/L	10	Standard
Fe	54	5879.5	1.7	**5.3975**	0.172	3.2	mg/L	179	Standard
Fe	57	1785.1	7.3	**5.6694**	0.625	11.0	mg/L	335	Standard
> **Sc-1**	45	54972.9	1.9				mg/L	61029	Standard
Cl	35	21821.7	0.9				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	1470.1	4.5				ug/L	1547	Standard
P	31	16677.5	6.0				ug/L	15286	Standard
S	34	1955.1	4.5				ug/L	1572	Standard
Sr	88	73.3	14.2				ug/L	73	Standard
C	12	80.0	66.1				mg/L	63	Standard
N	14	6.7	86.6				mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.9	32.3				mg/L	6	Standard
Ho-1	165	3.3	86.6				mg/L	10	Standard
Er	166	16.7	124.9				mg/L	13	Standard
I	127	1668.4	28.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	100.945		
Al	27	98.569		
Sc	45			
Ti	47	100.224		
V	51	103.207		
Cr	52	100.781		
Cr	53			
Mn	55	106.711		
Co	59	107.591		
Ni	60	100.481		
Cu	65	101.375		
Zn	66	101.952		
> Ge	72		97.448	
As	75	102.823		
Se	82	104.807		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:02:27
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	102.509		
Ag	107	101.612		
Cd	111	103.231		
Cd	114			
> In	115		96.281	
Sn	118	103.076		
Sb	123	101.661		
Ba	135	100.916		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	103.817		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.814		
U	238	105.283		
> Bi	209		96.559	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:02:27
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 12:07:34
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74348.6	1.0				ug/L	74699	Standard
	Be	9	21.7	26.6	**0.0324**	0.007	20.7	ug/L	8	Standard
	Al	27	1840.1	21.7	**-0.0008**	0.003	358.1	ug/L	1493	Standard
	Sc	45	56396.5	2.8				ug/L	61029	Standard
	Ti	47	36.7	15.0	**-0.0098**	0.014	147.0	ug/L	35	Standard
	V	51	1326.6	5.0	**-0.0153**	0.008	51.0	ug/L	1435	Standard
	Cr	52	6702.5	2.0	**-0.0955**	0.015	16.1	ug/L	7635	Standard
	Cr	53	393.3	15.3	**0.1070**	0.067	63.0	ug/L	363	Standard
	Mn	55	742.7	5.5	**-0.0386**	0.005	13.8	ug/L	747	Standard
	Co	59	250.3	1.5	**0.0016**	0.000	18.8	ug/L	234	Standard
	Ni	60	195.0	9.2	**-0.0170**	0.010	58.7	ug/L	215	Standard
	Cu	65	108.7	1.4	**-0.0438**	0.000	1.1	ug/L	104	Standard
	Zn	66	377.3	7.3	**-0.1000**	0.024	24.2	ug/L	363	Standard
>	**Ge**	72	482419.6	0.7				ug/L	495142	Standard
	As	75	-16.3	122.1	**0.0447**	0.018	40.8	ug/L	-50	Standard
	Se	82	12.2	10.1	**-0.0240**	0.012	49.4	ug/L	14	Standard
	Se-1	77	52.7	12.2	**0.1177**	0.089	75.6	ug/L	42	Standard
>	**Ga**	71	61.7	33.8				mg/L	42	Standard
	Rb	85	30.0	50.0				ug/L	17	Standard
	Y	89	414790.7	2.9				ug/L	417098	Standard
>	**Rh**	103	1.7	173.2				ug/L	13	Standard
	Mo	98	135.9	9.5	**0.0469**	0.005	10.2	ug/L	11	Standard
	Ag	107	96.3	7.3	**0.0024**	0.001	45.2	ug/L	68	Standard
	Cd	111	7.4	43.0	**-0.0042**	0.002	37.1	mg/L	2	Standard
	Cd	114	24.8	51.6	**0.0029**	0.003	88.9	ug/L	12	Standard
>	**In**	115	568967.0	0.5				ug/L	587441	Standard
	Sn	118	628.3	16.1	**0.0164**	0.018	110.8	ug/L	497	Standard
	Sb	123	617.2	40.6	**0.1324**	0.056	42.3	ug/L	78	Standard
	Ba	135	30.7	16.1	**-0.0005**	0.003	469.7	ug/L	24	Standard
	Ce	140	23.3	68.9				ug/L	25	Standard
>	**Tb**	159	918183.5	1.2				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	30.3	14.9	**0.0020**	0.001	31.1	ug/L	4	Standard
	Tl	205	23.3	24.7	**0.0035**	0.001	22.2	ug/L	2	Standard
	Pb	206	259.3	0.6	**-0.0221**	0.001	4.2	ug/L	240	Standard
	Pb	207	198.7	1.6	**-0.0277**	0.001	4.0	ug/L	198	Standard
	Pb	208	894.0	2.8	**-0.0259**	0.002	7.7	ug/L	883	Standard
	U	238	15.0	35.3	**-0.0022**	0.001	35.0	ug/L	3	Standard
>	**Bi**	209	518580.7	1.8				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:09:51
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**20.6253**	35.715	173.2	mg/L	0	Standard
Mg	24	31.7	32.9	**-0.0207**	0.015	73.7	mg/L	37	Standard
K	39	51.7	24.4	**0.0090**	0.027	299.1	mg/L	38	Standard
Ca	43	33.3	31.2	**-0.2319**	0.781	336.6	mg/L	10	Standard
Fe	54	140.8	7.6	**-0.0093**	0.006	67.4	mg/L	179	Standard
Fe	57	256.7	18.9	**-0.1295**	0.180	138.6	mg/L	335	Standard
Sc-1	45	56396.5	2.8				mg/L	61029	Standard
Cl	35	21372.4	1.3				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	1540.1	8.8				ug/L	1547	Standard
P	31	15931.7	0.9				ug/L	15286	Standard
S	34	1863.4	1.4				ug/L	1572	Standard
Sr	88	58.3	17.8				ug/L	73	Standard
C	12	56.7	27.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	6.2	186.6				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	10.0					mg/L	13	Standard
I	127	1518.4	26.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		97.431	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:09:51
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.855
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.969
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:09:51
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066106
Sample Date/Time: Wednesday, June 29, 2016 12:12:15
Number of Replicates: 3
Autosampler Position: 301
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78301.9	3.0				ug/L	74699	Standard
	Be	9	591.7	15.8	**0.6482**	0.094	14.5	ug/L	8	Standard
	Al	27	14033022.6	1.2	**91.1117**	1.566	1.7	ug/L	1493	Standard
	Sc	45	64031.7	1.2				ug/L	61029	Standard
	Ti	47	6936.9	1.2	**18.7002**	0.397	2.1	ug/L	35	Standard
	V	51	136379.5	0.6	**18.3618**	0.342	1.9	ug/L	1435	Standard
	Cr	52	135090.6	0.2	**17.7119**	0.504	2.8	ug/L	7635	Standard
	Cr	53	16100.2	3.4	**17.5840**	1.043	5.9	ug/L	363	Standard
	Mn	55	2727656.7	0.9	**402.5207**	10.967	2.7	ug/L	747	Standard
	Co	59	46530.0	0.2	**7.2050**	0.195	2.7	ug/L	234	Standard
	Ni	60	31243.2	1.0	**17.0295**	0.585	3.4	ug/L	215	Standard
	Cu	65	40278.5	0.5	**22.1396**	0.604	2.7	ug/L	104	Standard
	Zn	66	62615.1	0.3	**58.4626**	1.558	2.7	ug/L	363	Standard
>	**Ge**	72	481946.2	2.4				ug/L	495142	Standard
	As	75	4644.7	1.4	**4.3222**	0.154	3.6	ug/L	-50	Standard
	Se	82	54.0	14.6	**0.3688**	0.086	23.4	ug/L	14	Standard
	Se-1	77	96.0	5.2	**0.7275**	0.040	5.5	ug/L	42	Standard
>	**Ga**	71	12810.4	4.6				mg/L	42	Standard
	Rb	85	116256.0	0.7				ug/L	17	Standard
	Y	89	585719.0	1.0				ug/L	417098	Standard
>	**Rh**	103	45.0	33.3				ug/L	13	Standard
	Mo	98	935.2	3.6	**0.3550**	0.010	2.8	ug/L	11	Standard
	Ag	107	931.4	1.2	**0.1340**	0.001	0.5	ug/L	68	Standard
	Cd	111	782.7	0.6	**0.3771**	0.007	1.8	mg/L	2	Standard
	Cd	114	1959.7	7.4	**0.3906**	0.025	6.5	ug/L	12	Standard
>	**In**	115	571359.2	1.2				ug/L	587441	Standard
	Sn	118	1011.7	7.7	**0.0833**	0.012	14.7	ug/L	497	Standard
	Sb	123	336.5	28.9	**0.0698**	0.021	29.4	ug/L	78	Standard
	Ba	135	394959.7	1.6	**198.3814**	1.476	0.7	ug/L	24	Standard
	Ce	140	893729.8	1.7				ug/L	25	Standard
>	**Tb**	159	948738.0	1.3				ug/L	939701	Standard
	Ho	165	15724.8	2.6				ug/L	10	Standard
	Tl	203	658.0	2.7	**0.0791**	0.002	2.2	ug/L	4	Standard
	Tl	205	583.3	11.6	**0.0828**	0.010	12.2	ug/L	2	Standard
	Pb	206	370590.1	1.2	**70.2116**	0.816	1.2	ug/L	240	Standard
	Pb	207	304396.4	0.8	**64.9093**	0.988	1.5	ug/L	198	Standard
	Pb	208	1254498.7	1.7	**67.7133**	1.716	2.5	ug/L	883	Standard
	U	238	3382.4	2.2	**0.4935**	0.016	3.3	ug/L	3	Standard
>	**Bi**	209	517229.2	1.1				ug/L	523984	Standard

Sample ID: L1606066106
Report Date/Time: Wednesday, June 29, 2016 12:14:32
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**36.6019**	63.388	173.2	mg/L	0	Standard
Mg	24	128.3	15.7	**0.0927**	0.024	25.5	mg/L	37	Standard
K	39	566.7	14.3	**1.0807**	0.180	16.7	mg/L	38	Standard
Ca	43	418.3	12.7	**22.6893**	3.491	15.4	mg/L	10	Standard
Fe	54	12793.3	3.0	**10.1983**	0.198	1.9	mg/L	179	Standard
Fe	57	3678.8	1.9	**10.8461**	0.363	3.4	mg/L	335	Standard
> **Sc-1**	45	64031.7	1.2				mg/L	61029	Standard
Cl	35	22050.7	1.1				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2173.5	11.7				ug/L	1547	Standard
P	31	16252.1	5.4				ug/L	15286	Standard
S	34	2065.1	4.4				ug/L	1572	Standard
Sr	88	140.0	9.4				ug/L	73	Standard
C	12	143.3	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	203.3	10.2				mg/L	3	Standard
Dy	164	25060.4	3.8				mg/L	6	Standard
Ho-1	165	15724.8	2.6				mg/L	10	Standard
Er	166	13529.4	4.1				mg/L	13	Standard
I	127	8297.3	2.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		104.824	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.335	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606066106
Report Date/Time: Wednesday, June 29, 2016 12:14:32
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.262
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	98.711
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606066106
Report Date/Time: Wednesday, June 29, 2016 12:14:32
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066107
Sample Date/Time: Wednesday, June 29, 2016 12:15:26
Number of Replicates: 3
Autosampler Position: 302
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	79855.0	3.3				ug/L	74699	Standard
	Be	9	765.0	6.8	**0.8219**	0.077	9.4	ug/L	8	Standard
	Al	27	26230756.6	0.7	**167.0452**	4.333	2.6	ug/L	1493	Standard
	Sc	45	64585.8	4.1				ug/L	61029	Standard
	Ti	47	7966.1	1.8	**21.8574**	0.352	1.6	ug/L	35	Standard
	V	51	155637.7	1.4	**21.3440**	0.459	2.1	ug/L	1435	Standard
	Cr	52	137713.7	0.8	**18.3989**	0.313	1.7	ug/L	7635	Standard
	Cr	53	16806.0	3.9	**18.6780**	0.841	4.5	ug/L	363	Standard
	Mn	55	4726467.2	1.1	**709.3402**	2.876	0.4	ug/L	747	Standard
	Co	59	64550.2	0.1	**10.1794**	0.081	0.8	ug/L	234	Standard
	Ni	60	42709.8	0.1	**23.7183**	0.226	1.0	ug/L	215	Standard
	Cu	65	23683.8	1.0	**13.1970**	0.251	1.9	ug/L	104	Standard
	Zn	66	44795.8	1.0	**42.4101**	0.817	1.9	ug/L	363	Standard
>	Ge	72	473762.9	0.9				ug/L	495142	Standard
	As	75	7180.0	1.6	**6.7588**	0.066	1.0	ug/L	-50	Standard
	Se	82	142.2	5.2	**1.2157**	0.067	5.5	ug/L	14	Standard
	Se-1	77	110.3	7.6	**0.9569**	0.126	13.2	ug/L	42	Standard
>	Ga	71	17858.9	0.4				mg/L	42	Standard
	Rb	85	152815.2	1.7				ug/L	17	Standard
	Y	89	630725.5	1.5				ug/L	417098	Standard
>	Rh	103	41.7	25.0				ug/L	13	Standard
	Mo	98	1176.0	1.0	**0.4476**	0.010	2.2	ug/L	11	Standard
	Ag	107	471.0	1.7	**0.0613**	0.002	3.5	ug/L	68	Standard
	Cd	111	231.8	9.8	**0.1061**	0.013	12.1	mg/L	2	Standard
	Cd	114	548.2	12.6	**0.1078**	0.015	13.9	ug/L	12	Standard
>	In	115	571871.5	1.5				ug/L	587441	Standard
	Sn	118	1063.4	7.5	**0.0923**	0.015	16.1	ug/L	497	Standard
	Sb	123	198.5	41.2	**0.0396**	0.018	46.2	ug/L	78	Standard
	Ba	135	410562.2	0.2	**206.0643**	2.723	1.3	ug/L	24	Standard
	Ce	140	1020863.3	0.8				ug/L	25	Standard
>	Tb	159	935971.5	1.4				ug/L	939701	Standard
	Ho	165	18616.4	1.9				ug/L	10	Standard
	Tl	203	969.7	3.1	**0.1217**	0.006	4.6	ug/L	4	Standard
	Tl	205	781.7	5.1	**0.1148**	0.005	4.1	ug/L	2	Standard
	Pb	206	74583.8	0.8	**14.5751**	0.129	0.9	ug/L	240	Standard
	Pb	207	60293.4	1.5	**13.2552**	0.050	0.4	ug/L	198	Standard
	Pb	208	253434.0	0.7	**14.1059**	0.309	2.2	ug/L	883	Standard
	U	238	4763.7	0.8	**0.7216**	0.012	1.7	ug/L	3	Standard
>	Bi	209	499544.2	1.5				ug/L	523984	Standard

Sample ID: L1606066107
Report Date/Time: Wednesday, June 29, 2016 12:17:43
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**90.1824**	29.248	32.4	mg/L	0	Standard
Mg	24	323.3	10.9	**0.3294**	0.041	12.6	mg/L	37	Standard
K	39	788.4	10.3	**1.5298**	0.110	7.2	mg/L	38	Standard
Ca	43	663.3	6.3	**37.0621**	1.711	4.6	mg/L	10	Standard
Fe	54	18630.5	1.2	**14.8018**	0.561	3.8	mg/L	179	Standard
Fe	57	5110.9	3.8	**15.3724**	1.217	7.9	mg/L	335	Standard
Sc-1	45	64585.8	4.1				mg/L	61029	Standard
Cl	35	22235.6	2.9				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	23104.3	4.2				ug/L	1547	Standard
P	31	15829.9	2.1				ug/L	15286	Standard
S	34	2033.5	3.2				ug/L	1572	Standard
Sr	88	161.7	12.5				ug/L	73	Standard
C	12	96.7	41.8				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	31177.2	1.6				mg/L	6	Standard
Ho-1	165	18616.4	1.9				mg/L	10	Standard
Er	166	16976.2	1.7				mg/L	13	Standard
I	127	9242.8	4.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		106.903	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.682	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606066107
Report Date/Time: Wednesday, June 29, 2016 12:17:43
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.350
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.336
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606066107
Report Date/Time: Wednesday, June 29, 2016 12:17:43
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066108
Sample Date/Time: Wednesday, June 29, 2016 12:18:38
Number of Replicates: 3
Autosampler Position: 303
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77986.9	1.5				ug/L	74699	Standard
	Be	9	848.4	8.3	**0.9313**	0.087	9.4	ug/L	8	Standard
	Al	27	17300686.4	1.1	**112.7510**	0.475	0.4	ug/L	1493	Standard
	Sc	45	64965.7	2.0				ug/L	61029	Standard
	Ti	47	6993.9	0.6	**18.3803**	0.198	1.1	ug/L	35	Standard
	V	51	170039.3	0.3	**22.3624**	0.095	0.4	ug/L	1435	Standard
	Cr	52	135852.7	1.1	**17.3429**	0.173	1.0	ug/L	7635	Standard
	Cr	53	16400.5	1.1	**17.4496**	0.150	0.9	ug/L	363	Standard
	Mn	55	3732645.2	0.8	**537.0232**	5.661	1.1	ug/L	747	Standard
	Co	59	59499.9	1.1	**8.9910**	0.146	1.6	ug/L	234	Standard
	Ni	60	38977.4	1.0	**20.7345**	0.185	0.9	ug/L	215	Standard
	Cu	65	32324.4	0.5	**17.2976**	0.098	0.6	ug/L	104	Standard
	Zn	66	55484.7	0.8	**50.4392**	0.445	0.9	ug/L	363	Standard
>	**Ge**	72	494184.7	0.5				ug/L	495142	Standard
	As	75	5935.1	0.6	**5.3688**	0.057	1.1	ug/L	-50	Standard
	Se	82	56.5	10.0	**0.3782**	0.052	13.7	ug/L	14	Standard
	Se-1	77	118.0	0.8	**0.9965**	0.019	1.9	ug/L	42	Standard
>	**Ga**	71	16687.5	1.9				mg/L	42	Standard
	Rb	85	157851.4	1.2				ug/L	17	Standard
	Y	89	635672.9	1.6				ug/L	417098	Standard
>	**Rh**	103	68.3	27.7				ug/L	13	Standard
	Mo	98	1214.1	0.7	**0.4777**	0.002	0.4	ug/L	11	Standard
	Ag	107	767.0	3.0	**0.1120**	0.004	3.7	ug/L	68	Standard
	Cd	111	562.6	3.0	**0.2776**	0.008	3.0	mg/L	2	Standard
	Cd	114	1264.6	12.4	**0.2596**	0.032	12.4	ug/L	12	Standard
>	**In**	115	553515.9	0.3				ug/L	587441	Standard
	Sn	118	763.4	11.0	**0.0440**	0.015	34.7	ug/L	497	Standard
	Sb	123	242.9	15.0	**0.0510**	0.008	16.4	ug/L	78	Standard
	Ba	135	429181.6	1.6	**222.5293**	3.912	1.8	ug/L	24	Standard
	Ce	140	1102655.1	2.3				ug/L	25	Standard
>	**Tb**	159	923824.1	1.4				ug/L	939701	Standard
	Ho	165	18663.2	2.9				ug/L	10	Standard
	Tl	203	782.7	3.6	**0.0981**	0.005	5.3	ug/L	4	Standard
	Tl	205	701.7	8.6	**0.1034**	0.010	10.1	ug/L	2	Standard
	Pb	206	213507.8	0.5	**41.9534**	0.898	2.1	ug/L	240	Standard
	Pb	207	175202.2	0.5	**38.7416**	0.683	1.8	ug/L	198	Standard
	Pb	208	727362.3	0.6	**40.7140**	1.033	2.5	ug/L	883	Standard
	U	238	4175.6	1.0	**0.6335**	0.018	2.9	ug/L	3	Standard
>	**Bi**	209	498479.4	1.9				ug/L	523984	Standard

Sample ID: L1606066108
Report Date/Time: Wednesday, June 29, 2016 12:20:55
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.0364**	31.231	173.2	mg/L	0	Standard
Mg	24	140.0	21.7	**0.1047**	0.037	35.2	mg/L	37	Standard
K	39	708.4	12.8	**1.3550**	0.168	12.4	mg/L	38	Standard
Ca	43	508.3	5.0	**27.6291**	0.895	3.2	mg/L	10	Standard
Fe	54	15324.4	1.1	**12.0707**	0.309	2.6	mg/L	179	Standard
Fe	57	4262.3	3.3	**12.5338**	0.279	2.2	mg/L	335	Standard
> **Sc-1**	45	64965.7	2.0				mg/L	61029	Standard
Cl	35	21927.9	1.6				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	2823.6	3.3				ug/L	1547	Standard
P	31	16612.4	1.0				ug/L	15286	Standard
S	34	2085.1	7.7				ug/L	1572	Standard
Sr	88	245.0	14.1				ug/L	73	Standard
C	12	116.7	27.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	93.3	49.5				mg/L	3	Standard
Dy	164	30954.0	1.2				mg/L	6	Standard
Ho-1	165	18663.2	2.9				mg/L	10	Standard
Er	166	16615.8	2.6				mg/L	13	Standard
I	127	11766.2	5.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		104.402	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.807	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606066108
Report Date/Time: Wednesday, June 29, 2016 12:20:55
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	94.225
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.132
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606066108
Report Date/Time: Wednesday, June 29, 2016 12:20:55
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606066109
Sample Date/Time: Wednesday, June 29, 2016 12:21:49
Number of Replicates: 3
Autosampler Position: 304
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77037.0	1.5				ug/L	74699	Standard
	Be	9	378.3	6.2	**0.4242**	0.022	5.2	ug/L	8	Standard
	Al	27	13254869.5	1.3	**87.4672**	2.449	2.8	ug/L	1493	Standard
	Sc	45	63133.0	3.0				ug/L	61029	Standard
	Ti	47	8144.2	1.7	**21.2616**	0.854	4.0	ug/L	35	Standard
	V	51	109381.5	1.5	**14.2029**	0.301	2.1	ug/L	1435	Standard
	Cr	52	127693.7	2.3	**16.1132**	0.755	4.7	ug/L	7635	Standard
	Cr	53	15149.2	3.9	**15.9774**	1.000	6.3	ug/L	363	Standard
	Mn	55	3010101.1	0.9	**429.7771**	13.412	3.1	ug/L	747	Standard
	Co	59	32158.1	0.7	**4.8045**	0.098	2.0	ug/L	234	Standard
	Ni	60	23283.5	0.8	**12.2424**	0.383	3.1	ug/L	215	Standard
	Cu	65	9940.9	0.4	**5.2071**	0.131	2.5	ug/L	104	Standard
	Zn	66	24785.3	0.2	**22.1054**	0.470	2.1	ug/L	363	Standard
>	**Ge**	72	498160.0	2.3				ug/L	495142	Standard
	As	75	2943.4	0.7	**2.6727**	0.077	2.9	ug/L	-50	Standard
	Se	82	39.1	13.2	**0.2170**	0.054	24.7	ug/L	14	Standard
	Se-1	77	75.7	11.2	**0.4068**	0.103	25.3	ug/L	42	Standard
>	**Ga**	71	9034.4	2.0				mg/L	42	Standard
	Rb	85	68682.0	2.5				ug/L	17	Standard
	Y	89	539979.1	1.9				ug/L	417098	Standard
>	**Rh**	103	11.7	24.7				ug/L	13	Standard
	Mo	98	3058.1	1.5	**1.1809**	0.018	1.5	ug/L	11	Standard
	Ag	107	334.3	34.2	**0.0400**	0.017	43.3	ug/L	68	Standard
	Cd	111	137.9	39.4	**0.0601**	0.026	43.1	mg/L	2	Standard
	Cd	114	340.6	39.5	**0.0664**	0.026	39.4	ug/L	12	Standard
>	**In**	115	568023.6	1.3				ug/L	587441	Standard
	Sn	118	973.4	3.3	**0.0776**	0.003	4.4	ug/L	497	Standard
	Sb	123	217.2	33.7	**0.0438**	0.016	35.6	ug/L	78	Standard
	Ba	135	138215.3	1.3	**69.8220**	0.106	0.2	ug/L	24	Standard
	Ce	140	545871.5	1.0				ug/L	25	Standard
>	**Tb**	159	946980.1	0.3				ug/L	939701	Standard
	Ho	165	11547.7	1.8				ug/L	10	Standard
	Tl	203	458.7	30.8	**0.0539**	0.017	31.9	ug/L	4	Standard
	Tl	205	360.0	14.1	**0.0505**	0.007	14.1	ug/L	2	Standard
	Pb	206	30575.5	1.0	**5.6549**	0.063	1.1	ug/L	240	Standard
	Pb	207	24590.3	1.3	**5.1132**	0.071	1.4	ug/L	198	Standard
	Pb	208	101567.1	0.7	**5.3447**	0.038	0.7	ug/L	883	Standard
	U	238	5591.0	1.3	**0.8082**	0.011	1.4	ug/L	3	Standard
>	**Bi**	209	523773.4	0.1				ug/L	523984	Standard

Sample ID: L1606066109
Report Date/Time: Wednesday, June 29, 2016 12:24:06
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.1528**	33.165	173.2	mg/L	0	Standard
Mg	24	88.3	26.1	**0.0447**	0.026	57.5	mg/L	37	Standard
K	39	410.0	24.0	**0.7638**	0.217	28.4	mg/L	38	Standard
Ca	43	368.3	13.4	**20.0310**	3.643	18.2	mg/L	10	Standard
Fe	54	11096.4	1.9	**8.9581**	0.101	1.1	mg/L	179	Standard
Fe	57	2985.3	3.6	**8.7334**	0.110	1.3	mg/L	335	Standard
> **Sc-1**	45	63133.0	3.0				mg/L	61029	Standard
Cl	35	21859.8	1.0				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	1696.8	5.1				ug/L	1547	Standard
P	31	15684.8	2.6				ug/L	15286	Standard
S	34	1883.5	4.2				ug/L	1572	Standard
Sr	88	116.7	19.3				ug/L	73	Standard
C	12	100.0	26.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	18768.7	3.5				mg/L	6	Standard
Ho-1	165	11547.7	1.8				mg/L	10	Standard
Er	166	10790.5	1.6				mg/L	13	Standard
I	127	6703.2	5.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.130	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.609	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606066109
Report Date/Time: Wednesday, June 29, 2016 12:24:06
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	96.695
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	99.960
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606066109
Report Date/Time: Wednesday, June 29, 2016 12:24:06
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082203
Sample Date/Time: Wednesday, June 29, 2016 12:25:01
Number of Replicates: 3
Autosampler Position: 305
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75246.5	2.8				ug/L	74699	Standard
	Be	9	6.7	114.6	**0.0150**	0.008	55.4	ug/L	8	Standard
	Al	27	1140453.8	1.3	**7.6935**	0.186	2.4	ug/L	1493	Standard
	Sc	45	58880.9	1.2				ug/L	61029	Standard
	Ti	47	77.7	5.2	**0.0982**	0.010	10.0	ug/L	35	Standard
	V	51	1502.7	7.0	**0.0058**	0.017	290.5	ug/L	1435	Standard
	Cr	52	10404.9	0.3	**0.3964**	0.023	5.7	ug/L	7635	Standard
	Cr	53	878.4	11.3	**0.6323**	0.120	19.0	ug/L	363	Standard
	Mn	55	64288.8	1.0	**9.1920**	0.174	1.9	ug/L	747	Standard
	Co	59	1249.1	3.1	**0.1540**	0.006	4.1	ug/L	234	Standard
	Ni	60	673.3	4.1	**0.2398**	0.011	4.6	ug/L	215	Standard
	Cu	65	229.7	8.0	**0.0211**	0.010	48.0	ug/L	104	Standard
	Zn	66	2146.2	1.5	**1.5328**	0.038	2.5	ug/L	363	Standard
>	**Ge**	72	489582.8	1.5				ug/L	495142	Standard
	As	75	99.3	9.7	**0.1492**	0.008	5.6	ug/L	-50	Standard
	Se	82	43.5	22.3	**0.2617**	0.085	32.3	ug/L	14	Standard
	Se-1	77	67.7	6.2	**0.3148**	0.059	18.8	ug/L	42	Standard
>	**Ga**	71	61.7	40.8				mg/L	42	Standard
	Rb	85	88606.7	2.0				ug/L	17	Standard
	Y	89	430024.7	1.6				ug/L	417098	Standard
>	**Rh**	103	28.3	50.9				ug/L	13	Standard
	Mo	98	81.5	16.2	**0.0249**	0.005	20.1	ug/L	11	Standard
	Ag	107	97.7	18.1	**0.0021**	0.003	124.3	ug/L	68	Standard
	Cd	111	13.9	57.3	**-0.0013**	0.004	291.1	mg/L	2	Standard
	Cd	114	23.0	35.9	**0.0024**	0.002	67.9	ug/L	12	Standard
>	**In**	115	586341.4	0.7				ug/L	587441	Standard
	Sn	118	508.3	13.9	**-0.0076**	0.011	151.7	ug/L	497	Standard
	Sb	123	156.7	24.2	**0.0294**	0.008	27.7	ug/L	78	Standard
	Ba	135	76489.9	1.2	**37.4252**	0.302	0.8	ug/L	24	Standard
	Ce	140	665.0	13.3				ug/L	25	Standard
>	**Tb**	159	934158.4	1.1				ug/L	939701	Standard
	Ho	165	15.0	66.7				ug/L	10	Standard
	Tl	203	113.7	11.3	**0.0121**	0.002	13.7	ug/L	4	Standard
	Tl	205	93.3	44.6	**0.0133**	0.006	44.3	ug/L	2	Standard
	Pb	206	362.0	5.5	**-0.0029**	0.004	131.1	ug/L	240	Standard
	Pb	207	281.3	6.6	**-0.0103**	0.004	38.7	ug/L	198	Standard
	Pb	208	1197.0	7.1	**-0.0098**	0.005	49.2	ug/L	883	Standard
	U	238	25.0	34.9	**-0.0008**	0.001	168.5	ug/L	3	Standard
>	**Bi**	209	520173.1	0.7				ug/L	523984	Standard

Sample ID: L1606082203
Report Date/Time: Wednesday, June 29, 2016 12:27:18
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	3547.1	5.1	**4.6793**	0.186	4.0	mg/L	37	Standard
K	39	1096.7	1.1	**2.3987**	0.055	2.3	mg/L	38	Standard
Ca	43	51.7	27.9	**0.8562**	0.931	108.7	mg/L	10	Standard
Fe	54	433.8	1.2	**0.2430**	0.009	3.5	mg/L	179	Standard
Fe	57	376.7	13.8	**0.2532**	0.189	74.5	mg/L	335	Standard
Sc-1	45	58880.9	1.2				mg/L	61029	Standard
Cl	35	22906.6	0.4				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	9496.3	3.0				ug/L	1547	Standard
P	31	16589.1	2.9				ug/L	15286	Standard
S	34	2071.8	1.6				ug/L	1572	Standard
Sr	88	118.3	17.1				ug/L	73	Standard
C	12	53.3	10.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	19.0	47.5				mg/L	6	Standard
Ho-1	165	15.0	66.7				mg/L	10	Standard
Er	166	20.0	100.0				mg/L	13	Standard
I	127	14315.1	5.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		100.733	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		98.877	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606082203
Report Date/Time: Wednesday, June 29, 2016 12:27:18
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.813
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.273
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082302
Sample Date/Time: Wednesday, June 29, 2016 12:28:12
Number of Replicates: 3
Autosampler Position: 306
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75794.1	0.6				ug/L	74699	Standard
	Be	9	13.3	21.7	**0.0225**	0.003	14.2	ug/L	8	Standard
	Al	27	2805598.0	0.7	**18.8025**	0.242	1.3	ug/L	1493	Standard
	Sc	45	59848.0	1.2				ug/L	61029	Standard
	Ti	47	42.3	20.1	**0.0017**	0.021	1272.3	ug/L	35	Standard
	V	51	1764.1	8.5	**0.0361**	0.018	49.9	ug/L	1435	Standard
	Cr	52	8282.9	3.5	**0.0853**	0.037	43.4	ug/L	7635	Standard
	Cr	53	906.7	14.2	**0.6437**	0.138	21.5	ug/L	363	Standard
	Mn	55	61274.7	1.5	**8.5848**	0.068	0.8	ug/L	747	Standard
	Co	59	3907.8	2.1	**0.5500**	0.015	2.7	ug/L	234	Standard
	Ni	60	654.7	3.9	**0.2231**	0.012	5.6	ug/L	215	Standard
	Cu	65	222.3	1.8	**0.0148**	0.001	8.1	ug/L	104	Standard
	Zn	66	2204.5	3.0	**1.5481**	0.054	3.5	ug/L	363	Standard
>	**Ge**	72	498979.3	0.8				ug/L	495142	Standard
	As	75	72.7	46.8	**0.1239**	0.030	24.0	ug/L	-50	Standard
	Se	82	52.0	6.1	**0.3317**	0.029	8.6	ug/L	14	Standard
	Se-1	77	105.0	8.3	**0.8037**	0.108	13.5	ug/L	42	Standard
>	**Ga**	71	56.7	13.5				mg/L	42	Standard
	Rb	85	343.3	5.1				ug/L	17	Standard
	Y	89	428520.0	1.9				ug/L	417098	Standard
>	**Rh**	103	38.3	27.2				ug/L	13	Standard
	Mo	98	38.4	23.3	**0.0089**	0.004	39.5	ug/L	11	Standard
	Ag	107	77.7	2.7	**-0.0008**	0.000	47.2	ug/L	68	Standard
	Cd	111	5.6	57.3	**-0.0052**	0.002	30.1	mg/L	2	Standard
	Cd	114	18.7	97.5	**0.0016**	0.004	227.0	ug/L	12	Standard
>	**In**	115	580343.3	0.7				ug/L	587441	Standard
	Sn	118	485.0	23.2	**-0.0106**	0.020	190.8	ug/L	497	Standard
	Sb	123	114.5	31.8	**0.0207**	0.008	39.1	ug/L	78	Standard
	Ba	135	150889.3	1.2	**74.6061**	0.512	0.7	ug/L	24	Standard
	Ce	140	123.3	15.3				ug/L	25	Standard
>	**Tb**	159	936316.1	1.2				ug/L	939701	Standard
	Ho	165	18.3	31.5				ug/L	10	Standard
	Tl	203	189.0	9.5	**0.0214**	0.002	8.8	ug/L	4	Standard
	Tl	205	161.7	6.4	**0.0230**	0.002	7.6	ug/L	2	Standard
	Pb	206	327.3	6.0	**-0.0093**	0.004	47.2	ug/L	240	Standard
	Pb	207	269.3	8.1	**-0.0127**	0.005	36.6	ug/L	198	Standard
	Pb	208	1102.3	4.7	**-0.0147**	0.003	21.5	ug/L	883	Standard
	U	238	62.7	10.4	**0.0048**	0.001	17.9	ug/L	3	Standard
>	**Bi**	209	519166.0	1.5				ug/L	523984	Standard

Sample ID: L1606082302
Report Date/Time: Wednesday, June 29, 2016 12:30:29
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	5325.9	0.2	**6.9463**	0.099	1.4	mg/L	37	Standard
K	39	75.0	37.1	**0.0544**	0.061	111.8	mg/L	38	Standard
Ca	43	86.7	3.3	**3.0583**	0.130	4.3	mg/L	10	Standard
Fe	54	170.8	11.8	**0.0094**	0.018	186.9	mg/L	179	Standard
Fe	57	311.7	10.3	**0.0068**	0.121	1775.6	mg/L	335	Standard
> **Sc-1**	45	59848.0	1.2				mg/L	61029	Standard
Cl	35	23747.3	1.3				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	11654.5	4.8				ug/L	1547	Standard
P	31	16051.8	4.0				ug/L	15286	Standard
S	34	1918.5	1.8				ug/L	1572	Standard
Sr	88	140.0	9.4				ug/L	73	Standard
C	12	53.3	39.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	12.2	186.8				mg/L	6	Standard
Ho-1	165	18.3	31.5				mg/L	10	Standard
Er	166	23.3	24.7				mg/L	13	Standard
I	127	13899.7	5.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.466	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.775	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082302
Report Date/Time: Wednesday, June 29, 2016 12:30:29
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		98.792
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.080
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082303
Sample Date/Time: Wednesday, June 29, 2016 12:31:24
Number of Replicates: 3
Autosampler Position: 307
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76244.7	1.6				ug/L	74699	Standard
	Be	9	10.0	50.0	**0.0187**	0.005	28.8	ug/L	8	Standard
	Al	27	1789441.8	2.2	**11.9153**	0.077	0.6	ug/L	1493	Standard
	Sc	45	61370.7	1.9				ug/L	61029	Standard
	Ti	47	59.7	11.8	**0.0448**	0.018	40.0	ug/L	35	Standard
	V	51	1976.3	1.1	**0.0601**	0.005	8.2	ug/L	1435	Standard
	Cr	52	8828.6	1.1	**0.1403**	0.016	11.5	ug/L	7635	Standard
	Cr	53	945.0	0.5	**0.6692**	0.006	1.0	ug/L	363	Standard
	Mn	55	180374.5	2.7	**25.1617**	0.439	1.7	ug/L	747	Standard
	Co	59	619.0	3.1	**0.0543**	0.004	7.0	ug/L	234	Standard
	Ni	60	405.7	6.0	**0.0878**	0.011	12.8	ug/L	215	Standard
	Cu	65	208.0	10.3	**0.0056**	0.012	222.0	ug/L	104	Standard
	Zn	66	6989.3	1.3	**5.7979**	0.130	2.2	ug/L	363	Standard
>	**Ge**	72	506777.4	1.1				ug/L	495142	Standard
	As	75	17.4	166.6	**0.0747**	0.025	33.6	ug/L	-50	Standard
	Se	82	43.0	5.9	**0.2447**	0.022	8.9	ug/L	14	Standard
	Se-1	77	89.0	9.0	**0.5677**	0.097	17.0	ug/L	42	Standard
>	**Ga**	71	55.0	18.2				mg/L	42	Standard
	Rb	85	226.7	20.8				ug/L	17	Standard
	Y	89	430230.4	1.6				ug/L	417098	Standard
>	**Rh**	103	23.3	12.4				ug/L	13	Standard
	Mo	98	43.4	8.2	**0.0107**	0.002	14.5	ug/L	11	Standard
	Ag	107	80.3	9.2	**-0.0005**	0.001	248.5	ug/L	68	Standard
	Cd	111	11.9	16.8	**-0.0022**	0.001	43.2	mg/L	2	Standard
	Cd	114	28.1	50.1	**0.0034**	0.003	82.8	ug/L	12	Standard
>	**In**	115	584282.7	1.3				ug/L	587441	Standard
	Sn	118	445.0	7.0	**-0.0181**	0.005	24.8	ug/L	497	Standard
	Sb	123	113.1	24.2	**0.0201**	0.006	28.5	ug/L	78	Standard
	Ba	135	83408.3	1.0	**40.9596**	0.601	1.5	ug/L	24	Standard
	Ce	140	301.7	22.6				ug/L	25	Standard
>	**Tb**	159	946714.8	0.6				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	182.7	3.3	**0.0205**	0.001	3.9	ug/L	4	Standard
	Tl	205	160.0	30.8	**0.0227**	0.007	30.9	ug/L	2	Standard
	Pb	206	288.0	1.3	**-0.0170**	0.001	5.3	ug/L	240	Standard
	Pb	207	270.7	6.0	**-0.0126**	0.003	25.4	ug/L	198	Standard
	Pb	208	1097.0	10.5	**-0.0153**	0.006	38.6	ug/L	883	Standard
	U	238	866.7	1.0	**0.1222**	0.001	0.8	ug/L	3	Standard
>	**Bi**	209	520921.7	0.5				ug/L	523984	Standard

Sample ID: L1606082303
Report Date/Time: Wednesday, June 29, 2016 12:33:40
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.3655**	33.533	173.2	mg/L	0	Standard
Mg	24	4915.8	1.0	**6.2466**	0.131	2.1	mg/L	37	Standard
K	39	50.0	20.0	**-0.0042**	0.023	541.6	mg/L	38	Standard
Ca	43	65.0	0.0	**1.5606**	0.079	5.0	mg/L	10	Standard
Fe	54	160.0	11.0	**-0.0035**	0.013	359.3	mg/L	179	Standard
Fe	57	281.7	9.1	**-0.1230**	0.068	55.3	mg/L	335	Standard
Sc-1	45	61370.7	1.9				mg/L	61029	Standard
Cl	35	23872.8	0.3				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	8675.8	8.1				ug/L	1547	Standard
P	31	16263.7	3.0				ug/L	15286	Standard
S	34	1945.1	6.6				ug/L	1572	Standard
Sr	88	130.0	24.0				ug/L	73	Standard
C	12	46.7	44.6				mg/L	63	Standard
N	14	6.7	86.6				mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	19.0	57.5				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	20.0	100.0				mg/L	13	Standard
I	127	19577.6	2.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		102.070	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		102.350	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606082303
Report Date/Time: Wednesday, June 29, 2016 12:33:40
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.462
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.415
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606082304
Sample Date/Time: Wednesday, June 29, 2016 12:34:35
Number of Replicates: 3
Autosampler Position: 308
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76966.7	2.1				ug/L	74699	Standard
	Be	9	8.3	91.7	**0.0169**	0.009	50.5	ug/L	8	Standard
	Al	27	441491.6	0.4	**2.9030**	0.060	2.1	ug/L	1493	Standard
	Sc	45	60155.9	2.7				ug/L	61029	Standard
	Ti	47	44.0	12.0	**0.0058**	0.013	219.4	ug/L	35	Standard
	V	51	1675.2	9.7	**0.0239**	0.019	79.8	ug/L	1435	Standard
	Cr	52	8137.2	4.7	**0.0630**	0.039	62.0	ug/L	7635	Standard
	Cr	53	611.7	11.7	**0.3248**	0.069	21.3	ug/L	363	Standard
	Mn	55	105254.7	1.8	**14.8182**	0.156	1.1	ug/L	747	Standard
	Co	59	259.7	5.8	**0.0017**	0.003	156.1	ug/L	234	Standard
	Ni	60	230.0	4.9	**-0.0022**	0.007	321.8	ug/L	215	Standard
	Cu	65	165.3	1.5	**-0.0157**	0.002	12.9	ug/L	104	Standard
	Zn	66	2212.5	1.5	**1.5507**	0.010	0.6	ug/L	363	Standard
>	**Ge**	72	500132.0	1.1				ug/L	495142	Standard
	As	75	21.8	38.4	**0.0789**	0.008	9.6	ug/L	-50	Standard
	Se	82	26.9	8.6	**0.1047**	0.023	22.2	ug/L	14	Standard
	Se-1	77	64.0	26.1	**0.2460**	0.230	93.4	ug/L	42	Standard
>	**Ga**	71	55.0	41.7				mg/L	42	Standard
	Rb	85	421.7	18.5				ug/L	17	Standard
	Y	89	429897.2	1.0				ug/L	417098	Standard
>	**Rh**	103	6.7	43.3				ug/L	13	Standard
	Mo	98	29.2	13.0	**0.0052**	0.001	25.4	ug/L	11	Standard
	Ag	107	87.7	6.3	**0.0005**	0.001	182.2	ug/L	68	Standard
	Cd	111	7.0	28.7	**-0.0046**	0.001	19.4	mg/L	2	Standard
	Cd	114	21.6	74.7	**0.0021**	0.003	150.8	ug/L	12	Standard
>	**In**	115	589054.0	1.8				ug/L	587441	Standard
	Sn	118	438.3	5.7	**-0.0199**	0.003	16.6	ug/L	497	Standard
	Sb	123	106.2	8.9	**0.0185**	0.002	12.5	ug/L	78	Standard
	Ba	135	25689.8	1.2	**12.5057**	0.379	3.0	ug/L	24	Standard
	Ce	140	65.0	26.6				ug/L	25	Standard
>	**Tb**	159	953435.2	1.7				ug/L	939701	Standard
	Ho	165	13.3	43.3				ug/L	10	Standard
	Tl	203	91.7	10.9	**0.0091**	0.001	12.8	ug/L	4	Standard
	Tl	205	101.7	11.4	**0.0141**	0.002	11.0	ug/L	2	Standard
	Pb	206	281.0	2.2	**-0.0197**	0.001	5.9	ug/L	240	Standard
	Pb	207	219.0	9.1	**-0.0248**	0.004	16.3	ug/L	198	Standard
	Pb	208	999.0	3.0	**-0.0219**	0.001	6.3	ug/L	883	Standard
	U	238	2.7	21.7	**-0.0040**	0.000	2.0	ug/L	3	Standard
>	**Bi**	209	535313.4	0.3				ug/L	523984	Standard

Sample ID: L1606082304
Report Date/Time: Wednesday, June 29, 2016 12:36:52
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	1101.7	8.8	**1.3766**	0.099	7.2	mg/L	37	Standard
K	39	50.0	26.5	**-0.0017**	0.031	1783.7	mg/L	38	Standard
Ca	43	43.3	17.6	**0.2622**	0.555	211.4	mg/L	10	Standard
Fe	54	1079.1	4.5	**0.7903**	0.045	5.7	mg/L	179	Standard
Fe	57	550.0	4.8	**0.8245**	0.134	16.2	mg/L	335	Standard
> **Sc-1**	45	60155.9	2.7				mg/L	61029	Standard
Cl	35	23751.9	0.9				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	3913.8	0.4				ug/L	1547	Standard
P	31	16552.4	4.0				ug/L	15286	Standard
S	34	1948.5	5.9				ug/L	1572	Standard
Sr	88	68.3	30.5				ug/L	73	Standard
C	12	63.3	36.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	19.7	135.8				mg/L	6	Standard
Ho-1	165	13.3	43.3				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	5207.6	8.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.036	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.008	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606082304
Report Date/Time: Wednesday, June 29, 2016 12:36:52
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	100.275
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.162
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 12:37:48
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	72653.8	1.6				ug/L	74699	Standard
	Be	9	42693.4	2.4	**49.8778**	1.581	3.2	ug/L	8	Standard
	Al	27	7018269.3	1.5	**49.0877**	0.211	0.4	ug/L	1493	Standard
	Sc	45	58700.2	1.7				ug/L	61029	Standard
	Ti	47	37360.7	1.5	**96.4941**	1.122	1.2	ug/L	35	Standard
	V	51	389361.0	0.6	**50.3345**	0.950	1.9	ug/L	1435	Standard
	Cr	52	377308.6	0.5	**48.8761**	0.835	1.7	ug/L	7635	Standard
	Cr	53	46063.3	3.3	**48.5325**	2.229	4.6	ug/L	363	Standard
	Mn	55	355991.7	2.4	**49.9570**	0.746	1.5	ug/L	747	Standard
	Co	59	341523.2	1.7	**50.6452**	0.528	1.0	ug/L	234	Standard
	Ni	60	94080.5	0.6	**49.1230**	0.490	1.0	ug/L	215	Standard
	Cu	65	95078.4	0.9	**49.9689**	1.178	2.4	ug/L	104	Standard
	Zn	66	56040.5	0.8	**49.8274**	0.722	1.4	ug/L	363	Standard
>	Ge	72	505267.2	1.6				ug/L	495142	Standard
	As	75	56871.7	1.3	**49.8175**	0.372	0.7	ug/L	-50	Standard
	Se	82	5593.2	2.1	**49.8004**	0.347	0.7	ug/L	14	Standard
	Se-1	77	3775.1	1.7	**50.0311**	0.568	1.1	ug/L	42	Standard
>	Ga	71	75.0	13.3				mg/L	42	Standard
	Rb	85	800.0	10.3				ug/L	17	Standard
	Y	89	408700.7	0.8				ug/L	417098	Standard
>	Rh	103	18.3	41.7				ug/L	13	Standard
	Mo	98	257910.1	1.9	**100.4657**	4.126	4.1	ug/L	11	Standard
	Ag	107	316989.6	1.3	**50.4692**	0.906	1.8	ug/L	68	Standard
	Cd	111	101914.0	1.0	**50.5948**	1.545	3.1	mg/L	2	Standard
	Cd	114	242699.9	2.4	**49.1311**	2.013	4.1	ug/L	12	Standard
>	In	115	566154.7	2.7				ug/L	587441	Standard
	Sn	118	281464.1	1.7	**49.9153**	1.113	2.2	ug/L	497	Standard
	Sb	123	227316.6	0.8	**50.5420**	1.534	3.0	ug/L	78	Standard
	Ba	135	98379.4	0.9	**49.8785**	1.233	2.5	ug/L	24	Standard
	Ce	140	113.3	20.8				ug/L	25	Standard
>	Tb	159	932106.1	1.1				ug/L	939701	Standard
	Ho	165	8.3	91.7				ug/L	10	Standard
	Tl	203	406761.4	0.1	**49.7003**	0.911	1.8	ug/L	4	Standard
	Tl	205	354916.5	2.0	**49.9771**	1.877	3.8	ug/L	2	Standard
	Pb	206	263562.6	0.6	**49.6234**	0.874	1.8	ug/L	240	Standard
	Pb	207	235278.4	0.8	**49.8612**	1.016	2.0	ug/L	198	Standard
	Pb	208	929662.3	0.8	**49.8673**	1.247	2.5	ug/L	883	Standard
	U	238	336021.4	1.5	**49.1730**	1.579	3.2	ug/L	3	Standard
>	Bi	209	520331.8	1.7				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:40:05
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**19.5945**	33.930	173.2	mg/L	0	Standard
Mg	24	3858.8	4.6	**5.1144**	0.246	4.8	mg/L	37	Standard
K	39	2055.1	3.6	**4.6098**	0.203	4.4	mg/L	38	Standard
Ca	43	130.0	23.4	**5.9979**	1.898	31.6	mg/L	10	Standard
Fe	54	5737.9	3.9	**4.9227**	0.278	5.7	mg/L	179	Standard
Fe	57	1583.4	3.8	**4.5245**	0.301	6.7	mg/L	335	Standard
Sc-1	45	58700.2	1.7				mg/L	61029	Standard
Cl	35	23099.6	1.9				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	1760.1	10.8				ug/L	1547	Standard
P	31	17002.9	0.9				ug/L	15286	Standard
S	34	2305.2	8.6				ug/L	1572	Standard
Sr	88	76.7	10.0				ug/L	73	Standard
C	12	60.0	33.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	9.0	191.6				mg/L	6	Standard
Ho-1	165	8.3	91.7				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	2030.1	31.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	99.756		
Al	27	98.175		
Sc	45			
Ti	47	96.494		
V	51	100.669		
Cr	52	97.752		
Cr	53			
Mn	55	99.914		
Co	59	101.290		
Ni	60	98.246		
Cu	65	99.938		
Zn	66	99.655		
Ge	72		102.045	
As	75	99.635		
Se	82	99.601		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:40:05
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	100.466		
Ag	107	100.938		
Cd	111	101.190		
Cd	114			
> In	115			96.376
Sn	118	99.831		
Sb	123	101.084		
Ba	135	99.757		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	99.401		
Tl	205			
Pb	206			
Pb	207			
Pb	208	99.735		
U	238	98.346		
> Bi	209			99.303
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 12:40:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 12:40:59
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	73745.7	1.0				ug/L	74699	Standard
	Be	9	45.0	86.8	**0.0596**	0.045	76.2	ug/L	8	Standard
	Al	27	10763.7	138.2	**0.0612**	0.103	168.8	ug/L	1493	Standard
	Sc	45	59794.4	0.5				ug/L	61029	Standard
	Ti	47	63.7	40.2	**0.0565**	0.070	123.9	ug/L	35	Standard
	V	51	1309.0	5.1	**-0.0256**	0.009	36.2	ug/L	1435	Standard
	Cr	52	7144.0	4.7	**-0.0780**	0.064	81.5	ug/L	7635	Standard
	Cr	53	391.7	5.2	**0.0859**	0.030	34.5	ug/L	363	Standard
	Mn	55	2203.3	114.3	**0.1667**	0.364	218.6	ug/L	747	Standard
	Co	59	375.0	36.8	**0.0187**	0.022	116.6	ug/L	234	Standard
	Ni	60	226.7	25.3	**-0.0048**	0.033	679.9	ug/L	215	Standard
	Cu	65	137.7	45.7	**-0.0307**	0.035	113.4	ug/L	104	Standard
	Zn	66	391.7	8.9	**-0.1025**	0.038	37.5	ug/L	363	Standard
>	**Ge**	72	505094.9	1.9				ug/L	495142	Standard
	As	75	-30.0	135.0	**0.0336**	0.035	105.0	ug/L	-50	Standard
	Se	82	18.1	21.6	**0.0240**	0.035	145.0	ug/L	14	Standard
	Se-1	77	51.7	16.7	**0.0700**	0.104	148.9	ug/L	42	Standard
>	**Ga**	71	55.0	27.3				mg/L	42	Standard
	Rb	85	30.0	72.6				ug/L	17	Standard
	Y	89	416389.1	0.5				ug/L	417098	Standard
>	**Rh**	103	18.3	56.8				ug/L	13	Standard
	Mo	98	292.6	24.5	**0.1060**	0.029	27.7	ug/L	11	Standard
	Ag	107	123.7	36.8	**0.0063**	0.007	108.0	ug/L	68	Standard
	Cd	111	19.1	60.5	**0.0013**	0.005	411.5	mg/L	2	Standard
	Cd	114	51.4	75.2	**0.0080**	0.007	93.3	ug/L	12	Standard
>	**In**	115	578912.4	1.7				ug/L	587441	Standard
	Sn	118	756.7	23.6	**0.0364**	0.029	79.3	ug/L	497	Standard
	Sb	123	391.4	30.7	**0.0806**	0.025	30.9	ug/L	78	Standard
	Ba	135	64.3	74.2	**0.0157**	0.023	147.1	ug/L	24	Standard
	Ce	140	98.3	142.4				ug/L	25	Standard
>	**Tb**	159	938659.9	2.1				ug/L	939701	Standard
	Ho	165	16.7	45.8				ug/L	10	Standard
	Tl	203	80.3	33.4	**0.0079**	0.003	39.1	ug/L	4	Standard
	Tl	205	263.3	142.0	**0.0362**	0.051	140.6	ug/L	2	Standard
	Pb	206	290.0	20.9	**-0.0175**	0.011	60.1	ug/L	240	Standard
	Pb	207	255.0	4.1	**-0.0168**	0.002	14.3	ug/L	198	Standard
	Pb	208	1806.4	80.0	**0.0207**	0.074	360.4	ug/L	883	Standard
	U	238	234.3	149.2	**0.0289**	0.049	171.5	ug/L	3	Standard
>	**Bi**	209	529175.1	1.6				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:43:16
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	48.3	47.8	**-0.0014**	0.030	2108.5	mg/L	37	Standard
K	39	68.3	15.2	**0.0400**	0.023	56.9	mg/L	38	Standard
Ca	43	21.7	74.2	**-1.1227**	1.047	93.2	mg/L	10	Standard
Fe	54	169.2	23.1	**0.0081**	0.034	414.8	mg/L	179	Standard
Fe	57	310.0	16.4	**0.0019**	0.182	9789.6	mg/L	335	Standard
> **Sc-1**	45	59794.4	0.5				mg/L	61029	Standard
Cl	35	23378.7	0.4				ug/L	20195	Standard
Kr	83	4.3	58.1				ug/L	4	Standard
Br	81	1630.1	3.1				ug/L	1547	Standard
P	31	16767.6	2.6				ug/L	15286	Standard
S	34	2153.5	8.8				ug/L	1572	Standard
Sr	88	65.0	27.7				ug/L	73	Standard
C	12	66.7	17.3				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	9.4	186.5				mg/L	6	Standard
Ho-1	165	16.7	45.8				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	1680.1	29.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.010	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:43:16
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	98.548	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	100.991	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 12:43:16
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Wednesday, June 29, 2016 12:44:12
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	71218.7	0.8				ug/L	74699	Standard
	Be	9	178.3	13.3	**0.2201**	0.029	13.2	ug/L	8	Standard
	Al	27	1766.8	29.5	**-0.0008**	0.004	458.8	ug/L	1493	Standard
	Sc	45	56025.1	1.1				ug/L	61029	Standard
	Ti	47	28.0	7.1	**-0.0318**	0.007	20.5	ug/L	35	Standard
	V	51	4079.9	5.1	**0.3696**	0.041	11.2	ug/L	1435	Standard
	Cr	52	12699.0	0.7	**0.7676**	0.041	5.4	ug/L	7635	Standard
	Cr	53	1050.0	6.6	**0.8579**	0.085	9.9	ug/L	363	Standard
	Mn	55	4110.2	0.9	**0.4692**	0.021	4.5	ug/L	747	Standard
	Co	59	2790.3	0.6	**0.4045**	0.014	3.5	ug/L	234	Standard
	Ni	60	2962.3	1.6	**1.5299**	0.023	1.5	ug/L	215	Standard
	Cu	65	1521.1	1.5	**0.7509**	0.034	4.5	ug/L	104	Standard
	Zn	66	7672.3	2.1	**6.8861**	0.035	0.5	ug/L	363	Standard
>	**Ge**	72	473802.8	2.6				ug/L	495142	Standard
	As	75	358.5	4.9	**0.3943**	0.021	5.3	ug/L	-50	Standard
	Se	82	48.3	16.1	**0.3230**	0.086	26.5	ug/L	14	Standard
	Se-1	77	78.0	9.0	**0.4927**	0.075	15.2	ug/L	42	Standard
>	**Ga**	71	35.0	24.7				mg/L	42	Standard
	Rb	85	21.7	35.3				ug/L	17	Standard
	Y	89	406616.3	2.6				ug/L	417098	Standard
>	**Rh**	103	13.3	43.3				ug/L	13	Standard
	Mo	98	80.8	10.7	**0.0260**	0.003	11.1	ug/L	11	Standard
	Ag	107	2461.5	2.7	**0.3827**	0.004	1.0	ug/L	68	Standard
	Cd	111	469.2	3.6	**0.2272**	0.006	2.5	mg/L	2	Standard
	Cd	114	1121.7	10.1	**0.2273**	0.026	11.5	ug/L	12	Standard
>	**In**	115	560717.1	1.7				ug/L	587441	Standard
	Sn	118	496.7	12.6	**-0.0055**	0.012	217.4	ug/L	497	Standard
	Sb	123	1804.7	2.2	**0.4010**	0.016	3.9	ug/L	78	Standard
	Ba	135	1433.1	2.6	**0.7176**	0.020	2.8	ug/L	24	Standard
	Ce	140	53.3	43.3				ug/L	25	Standard
>	**Tb**	159	901191.9	2.3				ug/L	939701	Standard
	Ho	165	5.0	0.0				ug/L	10	Standard
	Tl	203	658.3	6.7	**0.0809**	0.005	6.0	ug/L	4	Standard
	Tl	205	580.0	6.7	**0.0841**	0.005	5.7	ug/L	2	Standard
	Pb	206	1314.1	4.2	**0.1836**	0.008	4.1	ug/L	240	Standard
	Pb	207	1090.4	5.0	**0.1679**	0.009	5.1	ug/L	198	Standard
	Pb	208	4577.3	1.3	**0.1788**	0.005	3.0	ug/L	883	Standard
	U	238	2625.2	2.7	**0.3906**	0.007	1.9	ug/L	3	Standard
>	**Bi**	209	505831.0	1.4				ug/L	523984	Standard

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 12:46:29
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	40.0	12.5	**-0.0088**	0.007	78.9	mg/L	37	Standard
K	39	38.3	19.9	**-0.0218**	0.018	84.6	mg/L	38	Standard
Ca	43	36.7	75.1	**-0.0119**	1.877	15817.5	mg/L	10	Standard
Fe	54	97.1	24.3	**-0.0484**	0.023	47.3	mg/L	179	Standard
Fe	57	295.0	11.9	**0.0175**	0.123	699.1	mg/L	335	Standard
Sc-1	45	56025.1	1.1				mg/L	61029	Standard
Cl	35	22767.8	1.3				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	1303.4	5.0				ug/L	1547	Standard
P	31	11956.4	1.3				ug/L	15286	Standard
S	34	2061.8	6.1				ug/L	1572	Standard
Sr	88	80.0	22.5				ug/L	73	Standard
C	12	80.0	33.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	0.0					mg/L	6	Standard
Ho-1	165	5.0	0.0				mg/L	10	Standard
Er	166	0.0					mg/L	13	Standard
I	127	1360.1	35.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	110.075		
Al	27			
Sc	45			
Ti	47			
V	51	92.400		
Cr	52	95.955		
Cr	53			
Mn	55	93.844		
Co	59	101.134		
Ni	60	95.616		
Cu	65	93.867		
Zn	66	110.178		
Ge	72		95.690	
As	75	98.586		
Se	82	80.738		
Se-1	77			
Ga	71			

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 12:46:29
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	95.685		
Cd	111	94.646		
Cd	114			
> In	115			95.451
Sn	118			
Sb	123	100.239		
Ba	135	95.683		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	101.186		
Tl	205			
Pb	206			
Pb	207			
Pb	208	89.410		
U	238	97.655		
> Bi	209			96.535
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 12:46:29
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBW 7 WG573548-02
Sample Date/Time: Wednesday, June 29, 2016 12:53:54
Number of Replicates: 3
Autosampler Position: 309
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	74857.8	1.6				ug/L	74699	Standard
	Be	9	16.7	75.5	**0.0264**	0.014	53.3	ug/L	8	Standard
	Al	27	80794.9	2.0	**0.5353**	0.014	2.5	ug/L	1493	Standard
	Sc	45	58461.0	3.0				ug/L	61029	Standard
	Ti	47	78.0	5.9	**0.0921**	0.013	14.5	ug/L	35	Standard
	V	51	1804.8	0.8	**0.0380**	0.004	11.0	ug/L	1435	Standard
	Cr	52	9792.5	0.9	**0.2676**	0.021	7.7	ug/L	7635	Standard
	Cr	53	633.3	5.7	**0.3397**	0.038	11.1	ug/L	363	Standard
	Mn	55	7868.0	0.9	**0.9563**	0.019	2.0	ug/L	747	Standard
	Co	59	389.7	6.7	**0.0204**	0.003	16.2	ug/L	234	Standard
	Ni	60	396.3	4.2	**0.0829**	0.008	9.4	ug/L	215	Standard
	Cu	65	182.0	8.2	**-0.0081**	0.009	109.2	ug/L	104	Standard
	Zn	66	1432.4	5.9	**0.8274**	0.087	10.5	ug/L	363	Standard
>	**Ge**	72	506709.7	1.0				ug/L	495142	Standard
	As	75	61.7	54.0	**0.1136**	0.029	25.8	ug/L	-50	Standard
	Se	82	57.7	7.0	**0.3762**	0.040	10.6	ug/L	14	Standard
	Se-1	77	48.3	5.2	**0.0243**	0.031	128.3	ug/L	42	Standard
>	**Ga**	71	106.7	35.5				mg/L	42	Standard
	Rb	85	413.3	4.9				ug/L	17	Standard
	Y	89	414741.4	0.4				ug/L	417098	Standard
>	**Rh**	103	15.0	100.0				ug/L	13	Standard
	Mo	98	46.5	23.9	**0.0125**	0.004	33.5	ug/L	11	Standard
	Ag	107	1549.7	3.3	**0.2354**	0.012	5.1	ug/L	68	Standard
	Cd	111	17.2	8.9	**0.0007**	0.001	113.7	mg/L	2	Standard
	Cd	114	63.6	36.2	**0.0109**	0.005	44.4	ug/L	12	Standard
>	**In**	115	562869.1	1.7				ug/L	587441	Standard
	Sn	118	736.7	11.6	**0.0370**	0.017	44.6	ug/L	497	Standard
	Sb	123	216.9	10.9	**0.0443**	0.006	12.6	ug/L	78	Standard
	Ba	135	906.0	5.8	**0.4461**	0.030	6.6	ug/L	24	Standard
	Ce	140	1735.1	5.3				ug/L	25	Standard
>	**Tb**	159	918530.3	1.8				ug/L	939701	Standard
	Ho	165	31.7	32.9				ug/L	10	Standard
	Tl	203	89.3	15.5	**0.0092**	0.002	18.3	ug/L	4	Standard
	Tl	205	86.7	13.3	**0.0124**	0.002	13.0	ug/L	2	Standard
	Pb	206	378.0	4.1	**0.0001**	0.003	2503.4	ug/L	240	Standard
	Pb	207	310.7	3.7	**-0.0040**	0.002	56.8	ug/L	198	Standard
	Pb	208	1268.7	1.9	**-0.0059**	0.001	22.1	ug/L	883	Standard
	U	238	52.0	18.9	**0.0032**	0.001	44.2	ug/L	3	Standard
>	**Bi**	209	520017.2	0.7				ug/L	523984	Standard

Sample ID: PBW 7 WG573548-02
Report Date/Time: Wednesday, June 29, 2016 12:56:11
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	98.3	32.7	**0.0678**	0.045	67.0	mg/L	37	Standard
K	39	61.7	9.4	**0.0281**	0.012	42.4	mg/L	38	Standard
Ca	43	31.7	55.5	**-0.4154**	1.202	289.4	mg/L	10	Standard
Fe	54	153.7	8.4	**-0.0024**	0.007	308.5	mg/L	179	Standard
Fe	57	316.7	15.8	**0.0490**	0.173	353.8	mg/L	335	Standard
> **Sc-1**	45	58461.0	3.0				mg/L	61029	Standard
Cl	35	22819.8	1.0				ug/L	20195	Standard
Kr	83	3.7	31.5				ug/L	4	Standard
Br	81	12281.6	4.1				ug/L	1547	Standard
P	31	15094.2	2.5				ug/L	15286	Standard
S	34	1958.5	7.0				ug/L	1572	Standard
Sr	88	78.3	19.5				ug/L	73	Standard
C	12	96.7	21.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	48.1	42.1				mg/L	6	Standard
Ho-1	165	31.7	32.9				mg/L	10	Standard
Er	166	40.0	25.0				mg/L	13	Standard
I	127	2060.1	27.3				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		100.213	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		102.336	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBW 7 WG573548-02
Report Date/Time: Wednesday, June 29, 2016 12:56:11
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.817
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.243
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 7 WG573548-02
Report Date/Time: Wednesday, June 29, 2016 12:56:11
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSW 7 WG573548-03
Sample Date/Time: Wednesday, June 29, 2016 12:57:05
Number of Replicates: 3
Autosampler Position: 310
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	75779.1	1.4				ug/L	74699	Standard
	Be	9	44266.3	2.0	**49.5806**	1.413	2.8	ug/L	8	Standard
	Al	27	95277.1	2.2	**0.6259**	0.022	3.5	ug/L	1493	Standard
	Sc	45	60008.6	1.7				ug/L	61029	Standard
	Ti	47	61.3	19.5	**0.0533**	0.033	61.1	ug/L	35	Standard
	V	51	383994.5	1.5	**50.7641**	0.334	0.7	ug/L	1435	Standard
	Cr	52	391642.5	1.8	**51.9470**	0.996	1.9	ug/L	7635	Standard
	Cr	53	48176.5	1.3	**51.9174**	0.408	0.8	ug/L	363	Standard
	Mn	55	379234.4	0.1	**54.4505**	0.406	0.7	ug/L	747	Standard
	Co	59	337441.4	0.9	**51.1841**	0.650	1.3	ug/L	234	Standard
	Ni	60	96944.2	1.6	**51.7745**	0.652	1.3	ug/L	215	Standard
	Cu	65	97447.4	0.9	**52.3759**	0.338	0.6	ug/L	104	Standard
	Zn	66	58710.9	1.4	**53.4195**	0.683	1.3	ug/L	363	Standard
>	Ge	72	493989.7	0.8				ug/L	495142	Standard
	As	75	54420.0	2.0	**48.7560**	0.745	1.5	ug/L	-50	Standard
	Se	82	5499.8	2.1	**50.0913**	1.110	2.2	ug/L	14	Standard
	Se-1	77	3757.5	2.3	**50.9460**	1.277	2.5	ug/L	42	Standard
>	Ga	71	331.7	6.3				mg/L	42	Standard
	Rb	85	1905.1	7.7				ug/L	17	Standard
	Y	89	430193.5	2.2				ug/L	417098	Standard
>	Rh	103	21.7	13.3				ug/L	13	Standard
	Mo	98	200.3	5.0	**0.0686**	0.004	5.6	ug/L	11	Standard
	Ag	107	325181.2	0.5	**49.3255**	0.456	0.9	ug/L	68	Standard
	Cd	111	102382.7	0.7	**48.4138**	0.589	1.2	mg/L	2	Standard
	Cd	114	246010.3	1.2	**47.4297**	0.453	1.0	ug/L	12	Standard
>	In	115	594099.2	0.9				ug/L	587441	Standard
	Sn	118	733.4	7.5	**0.0295**	0.010	35.4	ug/L	497	Standard
	Sb	123	222790.7	0.6	**47.1842**	0.681	1.4	ug/L	78	Standard
	Ba	135	100456.3	0.7	**48.5175**	0.574	1.2	ug/L	24	Standard
	Ce	140	3297.0	2.9				ug/L	25	Standard
>	Tb	159	953975.0	2.3				ug/L	939701	Standard
	Ho	165	81.7	3.5				ug/L	10	Standard
	Tl	203	411383.7	1.0	**49.0691**	0.983	2.0	ug/L	4	Standard
	Tl	205	358822.0	2.1	**49.3118**	1.192	2.4	ug/L	2	Standard
	Pb	206	272876.4	0.6	**50.1549**	0.805	1.6	ug/L	240	Standard
	Pb	207	232281.8	1.2	**48.0517**	0.976	2.0	ug/L	198	Standard
	Pb	208	938885.2	1.2	**49.1601**	1.120	2.3	ug/L	883	Standard
	U	238	337957.9	0.3	**48.2703**	0.691	1.4	ug/L	3	Standard
>	Bi	209	532979.3	1.2				ug/L	523984	Standard

Sample ID: LCSW 7 WG573548-03
Report Date/Time: Wednesday, June 29, 2016 12:59:22
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	40.0	12.5	**-0.0126**	0.007	52.7	mg/L	37	Standard
K	39	58.3	21.6	**0.0173**	0.030	174.9	mg/L	38	Standard
Ca	43	60.0	16.7	**1.3313**	0.636	47.7	mg/L	10	Standard
Fe	54	427.3	5.7	**0.2300**	0.018	7.6	mg/L	179	Standard
Fe	57	393.3	4.8	**0.2864**	0.089	31.1	mg/L	335	Standard
> **Sc-1**	45	60008.6	1.7				mg/L	61029	Standard
Cl	35	22904.0	1.4				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	1933.5	11.4				ug/L	1547	Standard
P	31	15693.1	2.0				ug/L	15286	Standard
S	34	1913.5	5.4				ug/L	1572	Standard
Sr	88	85.0	21.2				ug/L	73	Standard
C	12	106.7	37.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	149.3	43.4				mg/L	6	Standard
Ho-1	165	81.7	3.5				mg/L	10	Standard
Er	166	83.3	45.4				mg/L	13	Standard
I	127	1690.1	32.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		101.446	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.767	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 7 WG573548-03
Report Date/Time: Wednesday, June 29, 2016 12:59:22
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.133
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.717
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098623 WG573548-01
Sample Date/Time: Wednesday, June 29, 2016 13:00:17
Number of Replicates: 3
Autosampler Position: 311
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85536.1	4.6				ug/L	74699	Standard
	Be	9	131.7	4.4	**0.1385**	0.009	6.7	ug/L	8	Standard
	Al	27	51229429.5	1.9	**304.6734**	8.454	2.8	ug/L	1493	Standard
	Sc	45	58156.4	0.7				ug/L	61029	Standard
	Ti	47	291.7	4.0	**0.6976**	0.027	3.9	ug/L	35	Standard
	V	51	2731.2	8.5	**0.1837**	0.035	19.3	ug/L	1435	Standard
	Cr	52	11798.6	2.2	**0.6453**	0.034	5.3	ug/L	7635	Standard
	Cr	53	1541.7	10.7	**1.4199**	0.200	14.1	ug/L	363	Standard
	Mn	55	1519614.4	1.3	**228.6645**	1.850	0.8	ug/L	747	Standard
	Co	59	1958.8	0.9	**0.2737**	0.000	0.1	ug/L	234	Standard
	Ni	60	5664.7	1.6	**3.0483**	0.071	2.3	ug/L	215	Standard
	Cu	65	19925.4	0.1	**11.1201**	0.082	0.7	ug/L	104	Standard
	Zn	66	6263.0	2.6	**5.5559**	0.126	2.3	ug/L	363	Standard
>	**Ge**	72	472304.7	0.8				ug/L	495142	Standard
	As	75	721.9	3.2	**0.7353**	0.022	3.1	ug/L	-50	Standard
	Se	82	400.4	3.0	**3.6870**	0.142	3.9	ug/L	14	Standard
	Se-1	77	187.0	7.2	**2.0610**	0.173	8.4	ug/L	42	Standard
>	**Ga**	71	100.0	8.7				mg/L	42	Standard
	Rb	85	5192.6	2.9				ug/L	17	Standard
	Y	89	424150.6	2.5				ug/L	417098	Standard
>	**Rh**	103	291.7	21.8				ug/L	13	Standard
	Mo	98	1037.1	3.0	**0.3904**	0.016	4.1	ug/L	11	Standard
	Ag	107	150.3	11.8	**0.0106**	0.003	25.5	ug/L	68	Standard
	Cd	111	86.5	9.0	**0.0342**	0.004	11.5	mg/L	2	Standard
	Cd	114	235.1	23.0	**0.0445**	0.010	23.2	ug/L	12	Standard
>	**In**	115	577173.5	1.1				ug/L	587441	Standard
	Sn	118	1115.0	2.8	**0.0996**	0.004	4.3	ug/L	497	Standard
	Sb	123	1725.9	23.2	**0.3715**	0.084	22.5	ug/L	78	Standard
	Ba	135	22344.1	0.8	**11.0965**	0.201	1.8	ug/L	24	Standard
	Ce	140	1543.4	4.9				ug/L	25	Standard
>	**Tb**	159	936227.5	1.2				ug/L	939701	Standard
	Ho	165	78.3	35.2				ug/L	10	Standard
	Tl	203	240.0	4.8	**0.0287**	0.001	5.1	ug/L	4	Standard
	Tl	205	171.7	9.4	**0.0253**	0.002	9.3	ug/L	2	Standard
	Pb	206	666.7	6.9	**0.0592**	0.009	15.2	ug/L	240	Standard
	Pb	207	558.7	0.1	**0.0529**	0.000	0.3	ug/L	198	Standard
	Pb	208	2267.7	0.3	**0.0522**	0.000	0.9	ug/L	883	Standard
	U	238	28225.3	0.9	**4.2785**	0.042	1.0	ug/L	3	Standard
>	**Bi**	209	501665.1	0.1				ug/L	523984	Standard

Sample ID: L1606098623 WG573548-01
Report Date/Time: Wednesday, June 29, 2016 13:02:34
Page 1

Approved: July 04, 2016

Na	23	13.3	43.3	**161.5718**	71.131	44.0	mg/L	0	Standard
Mg	24	24900.4	1.1	**33.6647**	0.149	0.4	mg/L	37	Standard
K	39	1233.4	1.4	**2.7469**	0.062	2.2	mg/L	38	Standard
Ca	43	1560.1	5.5	**100.8648**	4.911	4.9	mg/L	10	Standard
Fe	54	167.1	25.8	**0.0105**	0.039	372.7	mg/L	179	Standard
Fe	57	958.4	6.3	**2.3440**	0.212	9.0	mg/L	335	Standard
Sc-1	45	58156.4	0.7				mg/L	61029	Standard
Cl	35	40465.3	2.5				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	58241.9	4.1				ug/L	1547	Standard
P	31	17453.4	1.6				ug/L	15286	Standard
S	34	1810.1	8.9				ug/L	1572	Standard
Sr	88	1313.4	8.1				ug/L	73	Standard
C	12	273.3	15.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	3	Standard
Dy	164	89.3	16.3				mg/L	6	Standard
Ho-1	165	78.3	35.2				mg/L	10	Standard
Er	166	83.3	18.3				mg/L	13	Standard
I	127	135078.2	7.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		114.508	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.388	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098623 WG573548-01
Report Date/Time: Wednesday, June 29, 2016 13:02:34
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.252
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.740
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098623 WG573548-01
Report Date/Time: Wednesday, June 29, 2016 13:02:34
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098635S WG573548-04
Sample Date/Time: Wednesday, June 29, 2016 13:03:28
Number of Replicates: 3
Autosampler Position: 312
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96645.6	0.2				ug/L	74699	Standard
	Be	9	56332.9	2.6	**49.4619**	1.222	2.5	ug/L	8	Standard
	Al	27	56565131.5	1.0	**297.4859**	3.609	1.2	ug/L	1493	Standard
	Sc	45	58999.8	4.0				ug/L	61029	Standard
	Ti	47	369.0	11.5	**0.8968**	0.108	12.0	ug/L	35	Standard
	V	51	390242.6	0.5	**53.2266**	0.984	1.8	ug/L	1435	Standard
	Cr	52	371857.5	2.1	**50.8419**	0.408	0.8	ug/L	7635	Standard
	Cr	53	46974.4	2.3	**52.2114**	1.040	2.0	ug/L	363	Standard
	Mn	55	1851169.7	0.3	**274.7367**	3.654	1.3	ug/L	747	Standard
	Co	59	350457.1	0.5	**54.8326**	0.988	1.8	ug/L	234	Standard
	Ni	60	94372.4	2.0	**51.9777**	0.274	0.5	ug/L	215	Standard
	Cu	65	115580.4	2.6	**64.0828**	0.734	1.1	ug/L	104	Standard
	Zn	66	61118.4	0.5	**57.3902**	0.612	1.1	ug/L	363	Standard
>	**Ge**	72	478993.7	1.6				ug/L	495142	Standard
	As	75	59122.1	0.8	**54.6260**	0.454	0.8	ug/L	-50	Standard
	Se	82	6294.5	0.9	**59.1515**	0.578	1.0	ug/L	14	Standard
	Se-1	77	4101.2	2.7	**57.4162**	0.662	1.2	ug/L	42	Standard
>	**Ga**	71	100.0	18.0				mg/L	42	Standard
	Rb	85	5364.3	1.3				ug/L	17	Standard
	Y	89	423648.2	2.3				ug/L	417098	Standard
>	**Rh**	103	290.0	15.8				ug/L	13	Standard
	Mo	98	1231.1	6.2	**0.4537**	0.027	5.8	ug/L	11	Standard
	Ag	107	140053.7	1.3	**21.3673**	0.341	1.6	ug/L	68	Standard
	Cd	111	101619.5	1.1	**48.3438**	0.154	0.3	mg/L	2	Standard
	Cd	114	248812.6	0.3	**48.2667**	0.577	1.2	ug/L	12	Standard
>	**In**	115	590497.7	1.3				ug/L	587441	Standard
	Sn	118	1276.7	2.9	**0.1227**	0.007	5.6	ug/L	497	Standard
	Sb	123	230648.9	1.4	**49.1452**	0.703	1.4	ug/L	78	Standard
	Ba	135	121100.4	1.0	**58.8551**	1.255	2.1	ug/L	24	Standard
	Ce	140	1580.1	12.1				ug/L	25	Standard
>	**Tb**	159	967679.2	1.8				ug/L	939701	Standard
	Ho	165	70.0	12.4				ug/L	10	Standard
	Tl	203	411213.8	0.9	**50.0486**	0.144	0.3	ug/L	4	Standard
	Tl	205	371375.3	1.6	**52.0881**	1.342	2.6	ug/L	2	Standard
	Pb	206	271184.0	0.3	**50.8655**	0.565	1.1	ug/L	240	Standard
	Pb	207	230172.1	0.3	**48.5918**	0.740	1.5	ug/L	198	Standard
	Pb	208	934549.0	0.8	**49.9358**	0.874	1.8	ug/L	883	Standard
	U	238	386476.6	1.1	**56.3336**	1.024	1.8	ug/L	3	Standard
>	**Bi**	209	522269.6	1.2				ug/L	523984	Standard

Sample ID: L1606098635S WG573548-04
Report Date/Time: Wednesday, June 29, 2016 13:05:45
Page 1

Approved: July 04, 2016

Na	23	8.3	91.7	**99.7813**	89.869	90.1	mg/L	0	Standard
Mg	24	26446.4	1.6	**35.2703**	0.863	2.4	mg/L	37	Standard
K	39	1201.7	4.6	**2.6333**	0.063	2.4	mg/L	38	Standard
Ca	43	1648.4	4.4	**105.1755**	1.490	1.4	mg/L	10	Standard
Fe	54	207.2	35.2	**0.0431**	0.061	140.9	mg/L	179	Standard
Fe	57	976.7	6.7	**2.3580**	0.131	5.6	mg/L	335	Standard
> **Sc-1**	45	58999.8	4.0				mg/L	61029	Standard
Cl	35	42664.0	0.9				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	62512.1	2.5				ug/L	1547	Standard
P	31	18094.1	2.0				ug/L	15286	Standard
S	34	1855.1	3.1				ug/L	1572	Standard
Sr	88	1376.7	4.2				ug/L	73	Standard
C	12	206.7	24.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	73.5	28.0				mg/L	6	Standard
Ho-1	165	70.0	12.4				mg/L	10	Standard
Er	166	66.7	34.6				mg/L	13	Standard
I	127	146421.6	4.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		129.380	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.739	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098635S WG573548-04
Report Date/Time: Wednesday, June 29, 2016 13:05:45
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	100.520
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	99.673
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098635S WG573548-04
Report Date/Time: Wednesday, June 29, 2016 13:05:45
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098637SD WG573548-05
Sample Date/Time: Wednesday, June 29, 2016 13:06:39
Number of Replicates: 3
Autosampler Position: 313
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	99478.9	1.8				ug/L	74699	Standard
	Be	9	55809.3	1.7	**47.6121**	0.679	1.4	ug/L	8	Standard
	Al	27	57079276.5	2.7	**291.6498**	6.801	2.3	ug/L	1493	Standard
	Sc	45	56220.8	1.7				ug/L	61029	Standard
	Ti	47	357.7	29.4	**0.8728**	0.310	35.5	ug/L	35	Standard
	V	51	392795.1	0.3	**53.7087**	1.385	2.6	ug/L	1435	Standard
	Cr	52	364211.2	1.2	**49.9063**	1.054	2.1	ug/L	7635	Standard
	Cr	53	45092.0	2.6	**50.2220**	0.967	1.9	ug/L	363	Standard
	Mn	55	1864825.2	1.1	**277.4410**	6.534	2.4	ug/L	747	Standard
	Co	59	357216.6	1.3	**56.0166**	0.665	1.2	ug/L	234	Standard
	Ni	60	94064.4	0.7	**51.9434**	1.099	2.1	ug/L	215	Standard
	Cu	65	96872.1	1.3	**53.8290**	0.581	1.1	ug/L	104	Standard
	Zn	66	56756.1	1.1	**53.3877**	0.715	1.3	ug/L	363	Standard
>	**Ge**	72	477920.9	2.4				ug/L	495142	Standard
	As	75	58834.8	0.8	**54.4924**	0.918	1.7	ug/L	-50	Standard
	Se	82	6411.7	0.6	**60.4029**	1.057	1.7	ug/L	14	Standard
	Se-1	77	4156.6	3.0	**58.3446**	1.540	2.6	ug/L	42	Standard
>	**Ga**	71	126.7	16.4				mg/L	42	Standard
	Rb	85	5344.3	1.7				ug/L	17	Standard
	Y	89	425761.1	0.9				ug/L	417098	Standard
>	**Rh**	103	338.3	14.4				ug/L	13	Standard
	Mo	98	943.7	3.1	**0.3564**	0.013	3.7	ug/L	11	Standard
	Ag	107	126054.2	0.9	**19.7696**	0.296	1.5	ug/L	68	Standard
	Cd	111	100507.1	1.6	**49.1528**	0.361	0.7	mg/L	2	Standard
	Cd	114	240725.5	3.1	**47.9933**	0.862	1.8	ug/L	12	Standard
>	**In**	115	574417.4	1.5				ug/L	587441	Standard
	Sn	118	1058.4	4.2	**0.0905**	0.005	5.8	ug/L	497	Standard
	Sb	123	229849.5	0.5	**50.3475**	0.456	0.9	ug/L	78	Standard
	Ba	135	119199.1	0.7	**59.5474**	0.534	0.9	ug/L	24	Standard
	Ce	140	1411.7	10.5				ug/L	25	Standard
>	**Tb**	159	946510.5	1.0				ug/L	939701	Standard
	Ho	165	61.7	32.8				ug/L	10	Standard
	Tl	203	404314.5	1.0	**50.2968**	0.720	1.4	ug/L	4	Standard
	Tl	205	368989.1	0.6	**52.8952**	1.157	2.2	ug/L	2	Standard
	Pb	206	274418.5	1.5	**52.6043**	0.331	0.6	ug/L	240	Standard
	Pb	207	233408.5	1.0	**50.3596**	0.394	0.8	ug/L	198	Standard
	Pb	208	945417.2	0.6	**51.6312**	0.704	1.4	ug/L	883	Standard
	U	238	390395.6	1.2	**58.1569**	0.854	1.5	ug/L	3	Standard
>	**Bi**	209	511024.4	1.7				ug/L	523984	Standard

Sample ID: L1606098637SD WG573548-05
Report Date/Time: Wednesday, June 29, 2016 13:08:56
Page 1

Approved: July 04, 2016

Na	23	13.3	21.7	**167.0389**	37.375	22.4	mg/L	0	Standard
Mg	24	26820.4	2.0	**37.5161**	0.222	0.6	mg/L	37	Standard
K	39	1165.0	5.8	**2.6799**	0.118	4.4	mg/L	38	Standard
Ca	43	1491.7	10.9	**99.7885**	11.256	11.3	mg/L	10	Standard
Fe	54	157.1	11.5	**0.0062**	0.015	238.1	mg/L	179	Standard
Fe	57	1011.7	5.7	**2.6573**	0.164	6.2	mg/L	335	Standard
Sc-1	45	56220.8	1.7				mg/L	61029	Standard
Cl	35	42892.6	1.5				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	62984.0	1.3				ug/L	1547	Standard
P	31	17081.3	3.0				ug/L	15286	Standard
S	34	1805.1	2.7				ug/L	1572	Standard
Sr	88	1383.4	10.5				ug/L	73	Standard
C	12	180.0	14.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	124.6	33.0				mg/L	6	Standard
Ho-1	165	61.7	32.8				mg/L	10	Standard
Er	166	43.3	35.3				mg/L	13	Standard
I	127	140810.4	4.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		133.173	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.522	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098637SD WG573548-05
Report Date/Time: Wednesday, June 29, 2016 13:08:56
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.783
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	97.527
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098637SD WG573548-05
Report Date/Time: Wednesday, June 29, 2016 13:08:56
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098601
Sample Date/Time: Wednesday, June 29, 2016 13:09:51
Number of Replicates: 3
Autosampler Position: 314
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	102324.4	1.8				ug/L	74699	Standard
	Be	9	88.3	32.2	**0.0810**	0.024	30.0	ug/L	8	Standard
	Al	27	5114148.7	1.9	**25.3952**	0.605	2.4	ug/L	1493	Standard
	Sc	45	60191.0	1.9				ug/L	61029	Standard
	Ti	47	2779.9	0.6	**7.0936**	0.108	1.5	ug/L	35	Standard
	V	51	5111.4	5.6	**0.4688**	0.033	7.0	ug/L	1435	Standard
	Cr	52	13420.3	3.5	**0.7542**	0.051	6.8	ug/L	7635	Standard
	Cr	53	1458.4	3.3	**1.2196**	0.051	4.2	ug/L	363	Standard
	Mn	55	181173.8	0.3	**25.4046**	0.180	0.7	ug/L	747	Standard
	Co	59	2348.5	2.0	**0.3119**	0.007	2.1	ug/L	234	Standard
	Ni	60	2310.5	3.1	**1.0878**	0.031	2.9	ug/L	215	Standard
	Cu	65	2308.5	3.6	**1.1141**	0.034	3.0	ug/L	104	Standard
	Zn	66	22662.3	0.5	**19.9177**	0.090	0.4	ug/L	363	Standard
>	**Ge**	72	504254.4	0.9				ug/L	495142	Standard
	As	75	859.1	5.1	**0.8126**	0.036	4.5	ug/L	-50	Standard
	Se	82	56.8	14.9	**0.3708**	0.079	21.2	ug/L	14	Standard
	Se-1	77	87.3	17.2	**0.5507**	0.190	34.5	ug/L	42	Standard
>	**Ga**	71	526.7	8.3				mg/L	42	Standard
	Rb	85	16517.3	5.3				ug/L	17	Standard
	Y	89	430302.4	1.6				ug/L	417098	Standard
>	**Rh**	103	86.7	33.8				ug/L	13	Standard
	Mo	98	1898.6	1.4	**0.7136**	0.022	3.1	ug/L	11	Standard
	Ag	107	115.3	8.0	**0.0050**	0.001	29.0	ug/L	68	Standard
	Cd	111	60.8	33.2	**0.0215**	0.010	47.6	mg/L	2	Standard
	Cd	114	128.6	20.8	**0.0232**	0.005	20.9	ug/L	12	Standard
>	**In**	115	581907.5	1.7				ug/L	587441	Standard
	Sn	118	1231.7	10.8	**0.1183**	0.024	20.7	ug/L	497	Standard
	Sb	123	520.8	14.7	**0.1084**	0.016	15.1	ug/L	78	Standard
	Ba	135	41346.0	0.9	**20.3791**	0.217	1.1	ug/L	24	Standard
	Ce	140	23020.8	2.4				ug/L	25	Standard
>	**Tb**	159	964144.3	1.8				ug/L	939701	Standard
	Ho	165	513.3	9.3				ug/L	10	Standard
	Tl	203	211.3	12.7	**0.0234**	0.003	14.7	ug/L	4	Standard
	Tl	205	253.3	62.3	**0.0350**	0.022	63.0	ug/L	2	Standard
	Pb	206	1500.1	2.6	**0.2039**	0.008	4.0	ug/L	240	Standard
	Pb	207	1190.0	4.0	**0.1756**	0.010	5.7	ug/L	198	Standard
	Pb	208	4936.0	2.0	**0.1838**	0.008	4.3	ug/L	883	Standard
	U	238	2034.5	10.8	**0.2852**	0.034	11.8	ug/L	3	Standard
>	**Bi**	209	535012.0	1.1				ug/L	523984	Standard

Sample ID: L1606098601
Report Date/Time: Wednesday, June 29, 2016 13:12:08
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	50263.2	0.4	**65.7389**	1.477	2.2	mg/L	37	Standard
K	39	250.0	7.2	**0.4466**	0.047	10.5	mg/L	38	Standard
Ca	43	233.3	8.9	**12.4104**	1.127	9.1	mg/L	10	Standard
Fe	54	825.5	3.8	**0.5718**	0.038	6.7	mg/L	179	Standard
Fe	57	516.7	4.4	**0.7058**	0.045	6.4	mg/L	335	Standard
> **Sc-1**	45	60191.0	1.9				mg/L	61029	Standard
Cl	35	28383.9	2.7				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	8439.0	2.4				ug/L	1547	Standard
P	31	17857.2	2.8				ug/L	15286	Standard
S	34	1791.8	4.6				ug/L	1572	Standard
Sr	88	506.7	4.7				ug/L	73	Standard
C	12	153.3	22.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	893.4	9.1				mg/L	6	Standard
Ho-1	165	513.3	9.3				mg/L	10	Standard
Er	166	416.7	13.2				mg/L	13	Standard
I	127	18778.3	7.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		136.983	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.840	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098601
Report Date/Time: Wednesday, June 29, 2016 13:12:08
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	99.058	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	102.105	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603
Sample Date/Time: Wednesday, June 29, 2016 13:13:02
Number of Replicates: 3
Autosampler Position: 315
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	106961.3	1.6				ug/L	74699	Standard
	Be	9	178.3	7.1	**0.1492**	0.012	8.2	ug/L	8	Standard
	Al	27	64490531.4	4.4	**306.3886**	9.864	3.2	ug/L	1493	Standard
	Sc	45	59811.2	2.1				ug/L	61029	Standard
	Ti	47	416.0	0.2	**0.9910**	0.025	2.5	ug/L	35	Standard
	V	51	3004.3	2.1	**0.2029**	0.003	1.2	ug/L	1435	Standard
	Cr	52	11901.7	1.5	**0.5854**	0.019	3.3	ug/L	7635	Standard
	Cr	53	991.7	2.3	**0.7459**	0.050	6.7	ug/L	363	Standard
	Mn	55	12370472.2	4.2	**1779.3602**	31.656	1.8	ug/L	747	Standard
	Co	59	18875.4	5.7	**2.8249**	0.100	3.5	ug/L	234	Standard
	Ni	60	12202.3	4.0	**6.4037**	0.114	1.8	ug/L	215	Standard
	Cu	65	16836.0	2.8	**8.9582**	0.039	0.4	ug/L	104	Standard
	Zn	66	24869.4	2.3	**22.3549**	0.039	0.2	ug/L	363	Standard
>	**Ge**	72	494237.2	2.4				ug/L	495142	Standard
	As	75	1554.4	6.7	**1.4489**	0.061	4.2	ug/L	-50	Standard
	Se	82	71.5	7.7	**0.5142**	0.037	7.1	ug/L	14	Standard
	Se-1	77	63.0	12.4	**0.2431**	0.119	49.1	ug/L	42	Standard
>	**Ga**	71	356.7	10.3				mg/L	42	Standard
	Rb	85	9309.5	2.8				ug/L	17	Standard
	Y	89	460268.8	4.1				ug/L	417098	Standard
>	**Rh**	103	310.0	9.7				ug/L	13	Standard
	Mo	98	1407.8	3.5	**0.5012**	0.005	1.0	ug/L	11	Standard
	Ag	107	101.3	22.9	**0.0020**	0.003	151.7	ug/L	68	Standard
	Cd	111	246.3	7.9	**0.1051**	0.008	7.2	mg/L	2	Standard
	Cd	114	605.6	11.0	**0.1111**	0.009	8.4	ug/L	12	Standard
>	**In**	115	612194.6	4.4				ug/L	587441	Standard
	Sn	118	1470.1	15.0	**0.1461**	0.027	18.7	ug/L	497	Standard
	Sb	123	727.1	5.4	**0.1452**	0.003	2.2	ug/L	78	Standard
	Ba	135	24533.2	4.5	**11.4855**	0.055	0.5	ug/L	24	Standard
	Ce	140	44839.6	2.7				ug/L	25	Standard
>	**Tb**	159	1002318.5	3.7				ug/L	939701	Standard
	Ho	165	778.4	7.8				ug/L	10	Standard
	Tl	203	123.7	14.4	**0.0129**	0.002	13.5	ug/L	4	Standard
	Tl	205	116.7	9.9	**0.0161**	0.001	9.2	ug/L	2	Standard
	Pb	206	819.0	5.9	**0.0785**	0.008	10.4	ug/L	240	Standard
	Pb	207	658.0	2.8	**0.0655**	0.007	10.9	ug/L	198	Standard
	Pb	208	2733.4	3.4	**0.0682**	0.002	2.2	ug/L	883	Standard
	U	238	29463.0	4.0	**4.1728**	0.071	1.7	ug/L	3	Standard
>	**Bi**	209	536825.9	2.9				ug/L	523984	Standard

Sample ID: L1606098603
Report Date/Time: Wednesday, June 29, 2016 13:15:19
Page 1

Approved: July 04, 2016

Na	23	18.3	31.5	**216.6166**	73.255	33.8	mg/L	0	Standard
Mg	24	7246.7	2.5	**9.4794**	0.084	0.9	mg/L	37	Standard
K	39	393.3	11.0	**0.7727**	0.091	11.8	mg/L	38	Standard
Ca	43	1938.5	3.9	**122.4130**	4.002	3.3	mg/L	10	Standard
Fe	54	1213.7	6.6	**0.9123**	0.078	8.6	mg/L	179	Standard
Fe	57	1360.1	3.5	**3.6434**	0.167	4.6	mg/L	335	Standard
> **Sc-1**	45	59811.2	2.1				mg/L	61029	Standard
Cl	35	41831.0	3.4				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	8062.1	5.5				ug/L	1547	Standard
P	31	19340.7	4.7				ug/L	15286	Standard
S	34	2026.8	7.6				ug/L	1572	Standard
Sr	88	1283.4	13.4				ug/L	73	Standard
C	12	153.3	19.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	1258.4	10.3				mg/L	6	Standard
Ho-1	165	778.4	7.8				mg/L	10	Standard
Er	166	800.0	9.9				mg/L	13	Standard
I	127	76815.4	9.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		143.190	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.817	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098603
Report Date/Time: Wednesday, June 29, 2016 13:15:19
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	104.214
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.451
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098603
Report Date/Time: Wednesday, June 29, 2016 13:15:19
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603PS WG574572-01
Sample Date/Time: Wednesday, June 29, 2016 13:16:14
Number of Replicates: 3
Autosampler Position: 316
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	107768.6	1.4				ug/L	74699	Standard
	Be	9	63154.7	2.1	**49.7293**	0.735	1.5	ug/L	8	Standard
	Al	27	62883924.9	0.4	**296.6156**	4.197	1.4	ug/L	1493	Standard
	Sc	45	60669.6	1.7				ug/L	61029	Standard
	Ti	47	431.3	11.1	**1.0313**	0.120	11.7	ug/L	35	Standard
	V	51	423481.5	0.7	**56.0150**	0.782	1.4	ug/L	1435	Standard
	Cr	52	394353.4	1.0	**52.3183**	0.613	1.2	ug/L	7635	Standard
	Cr	53	48385.5	1.4	**52.1545**	1.086	2.1	ug/L	363	Standard
	Mn	55	12676353.0	0.9	**1825.0939**	28.984	1.6	ug/L	747	Standard
	Co	59	406698.1	0.5	**61.7025**	0.702	1.1	ug/L	234	Standard
	Ni	60	105609.8	0.7	**56.4229**	0.793	1.4	ug/L	215	Standard
	Cu	65	109624.8	0.4	**58.9403**	0.355	0.6	ug/L	104	Standard
	Zn	66	80974.1	1.0	**73.8592**	1.163	1.6	ug/L	363	Standard
>	Ge	72	493944.8	0.7				ug/L	495142	Standard
	As	75	62963.2	0.4	**56.4099**	0.584	1.0	ug/L	-50	Standard
	Se	82	6528.2	1.4	**59.4919**	1.250	2.1	ug/L	14	Standard
	Se-1	77	4208.9	2.0	**57.1456**	1.224	2.1	ug/L	42	Standard
>	Ga	71	341.7	3.7				mg/L	42	Standard
	Rb	85	9437.9	2.3				ug/L	17	Standard
	Y	89	453637.5	2.0				ug/L	417098	Standard
>	Rh	103	333.3	4.8				ug/L	13	Standard
	Mo	98	1406.8	1.9	**0.5023**	0.008	1.6	ug/L	11	Standard
	Ag	107	336751.7	1.1	**49.7241**	0.838	1.7	ug/L	68	Standard
	Cd	111	110645.7	0.7	**50.9305**	0.669	1.3	mg/L	2	Standard
	Cd	114	268194.0	0.5	**50.3333**	0.438	0.9	ug/L	12	Standard
>	In	115	610323.2	0.7				ug/L	587441	Standard
	Sn	118	1538.4	2.3	**0.1588**	0.006	3.5	ug/L	497	Standard
	Sb	123	241381.6	1.1	**49.7617**	0.807	1.6	ug/L	78	Standard
	Ba	135	126394.4	1.4	**59.4274**	1.193	2.0	ug/L	24	Standard
	Ce	140	44334.8	0.8				ug/L	25	Standard
>	Tb	159	999714.8	1.0				ug/L	939701	Standard
	Ho	165	786.7	10.3				ug/L	10	Standard
	Tl	203	445077.8	1.6	**52.3048**	0.918	1.8	ug/L	4	Standard
	Tl	205	409895.5	2.2	**55.5017**	1.265	2.3	ug/L	2	Standard
	Pb	206	293321.0	0.4	**53.1218**	0.228	0.4	ug/L	240	Standard
	Pb	207	252054.7	0.9	**51.3779**	0.617	1.2	ug/L	198	Standard
	Pb	208	1026149.4	0.5	**52.9396**	0.195	0.4	ug/L	883	Standard
	U	238	420484.7	1.6	**59.1737**	0.984	1.7	ug/L	3	Standard
>	Bi	209	540895.8	0.3				ug/L	523984	Standard

Sample ID: L1606098603PS WG574572-01
Report Date/Time: Wednesday, June 29, 2016 13:18:31
Page 1

Approved: July 04, 2016

Na	23	16.7	45.8	**192.4294**	85.849	44.6	mg/L	0	Standard
Mg	24	7256.7	2.5	**9.3609**	0.349	3.7	mg/L	37	Standard
K	39	406.7	8.6	**0.7894**	0.064	8.1	mg/L	38	Standard
Ca	43	1853.4	4.6	**115.2974**	6.640	5.8	mg/L	10	Standard
Fe	54	1091.3	5.4	**0.7921**	0.036	4.5	mg/L	179	Standard
Fe	57	1315.1	6.5	**3.4198**	0.224	6.6	mg/L	335	Standard
> **Sc-1**	45	60669.6	1.7				mg/L	61029	Standard
Cl	35	43150.7	2.4				ug/L	20195	Standard
Kr	83	2.3	49.5				ug/L	4	Standard
Br	81	7932.1	5.6				ug/L	1547	Standard
P	31	18981.9	3.1				ug/L	15286	Standard
S	34	1910.1	1.6				ug/L	1572	Standard
Sr	88	1396.7	9.1				ug/L	73	Standard
C	12	166.7	22.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	1221.7	7.2				mg/L	6	Standard
Ho-1	165	786.7	10.3				mg/L	10	Standard
Er	166	800.0	7.6				mg/L	13	Standard
I	127	76722.5	5.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		144.271	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.758	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098603PS WG574572-01
Report Date/Time: Wednesday, June 29, 2016 13:18:31
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	103.895
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	103.227
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098603PS WG574572-01
Report Date/Time: Wednesday, June 29, 2016 13:18:31
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603SDL WG574572-02
Sample Date/Time: Wednesday, June 29, 2016 13:19:25
Number of Replicates: 3
Autosampler Position: 317
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	109627.0	2.2				ug/L	74699	Standard
	Be	9	55.0	32.8	**0.0501**	0.014	27.0	ug/L	8	Standard
	Al	27	13052757.9	2.1	**60.5215**	1.604	2.7	ug/L	1493	Standard
	Sc	45	62358.1	1.0				ug/L	61029	Standard
	Ti	47	123.0	14.7	**0.1926**	0.050	25.7	ug/L	35	Standard
	V	51	2451.6	1.0	**0.1057**	0.008	7.9	ug/L	1435	Standard
	Cr	52	10290.8	1.8	**0.2648**	0.008	2.8	ug/L	7635	Standard
	Cr	53	588.3	10.7	**0.2610**	0.063	24.3	ug/L	363	Standard
	Mn	55	2629853.6	1.6	**350.4744**	10.232	2.9	ug/L	747	Standard
	Co	59	4233.3	3.1	**0.5577**	0.018	3.3	ug/L	234	Standard
	Ni	60	2886.6	1.8	**1.3071**	0.037	2.8	ug/L	215	Standard
	Cu	65	3859.5	1.6	**1.8207**	0.021	1.2	ug/L	104	Standard
	Zn	66	7087.3	1.2	**5.5680**	0.153	2.7	ug/L	363	Standard
>	**Ge**	72	533577.2	1.8				ug/L	495142	Standard
	As	75	314.5	3.2	**0.3203**	0.010	3.2	ug/L	-50	Standard
	Se	82	28.2	16.0	**0.1004**	0.037	37.2	ug/L	14	Standard
	Se-1	77	51.7	9.9	**0.0349**	0.075	214.8	ug/L	42	Standard
>	**Ga**	71	153.3	21.0				mg/L	42	Standard
	Rb	85	1916.8	2.5				ug/L	17	Standard
	Y	89	457911.1	1.7				ug/L	417098	Standard
>	**Rh**	103	73.3	17.2				ug/L	13	Standard
	Mo	98	334.8	7.8	**0.1120**	0.009	8.5	ug/L	11	Standard
	Ag	107	142.0	14.1	**0.0075**	0.003	37.0	ug/L	68	Standard
	Cd	111	78.8	16.1	**0.0274**	0.006	20.4	mg/L	2	Standard
	Cd	114	171.7	29.4	**0.0293**	0.009	31.0	ug/L	12	Standard
>	**In**	115	626915.6	0.6				ug/L	587441	Standard
	Sn	118	618.3	15.8	**0.0044**	0.015	338.7	ug/L	497	Standard
	Sb	123	2031.6	21.7	**0.4032**	0.086	21.3	ug/L	78	Standard
	Ba	135	5054.8	0.4	**2.2983**	0.024	1.0	ug/L	24	Standard
	Ce	140	9302.9	0.9				ug/L	25	Standard
>	**Tb**	159	1014481.9	2.3				ug/L	939701	Standard
	Ho	165	175.0	2.9				ug/L	10	Standard
	Tl	203	125.0	17.4	**0.0120**	0.002	18.4	ug/L	4	Standard
	Tl	205	85.0	15.6	**0.0110**	0.002	15.7	ug/L	2	Standard
	Pb	206	501.3	9.2	**0.0141**	0.008	53.2	ug/L	240	Standard
	Pb	207	392.0	5.3	**0.0051**	0.005	98.5	ug/L	198	Standard
	Pb	208	1534.0	5.0	**0.0003**	0.005	1457.8	ug/L	883	Standard
	U	238	6033.2	1.4	**0.7925**	0.020	2.5	ug/L	3	Standard
>	**Bi**	209	576425.9	1.5				ug/L	523984	Standard

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:21:42
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**37.9583**	65.737	173.2	mg/L	0	Standard
Mg	24	1580.1	4.3	**1.9318**	0.105	5.4	mg/L	37	Standard
K	39	138.3	24.1	**0.1853**	0.073	39.3	mg/L	38	Standard
Ca	43	413.3	5.7	**23.0292**	1.446	6.3	mg/L	10	Standard
Fe	54	367.9	5.4	**0.1671**	0.019	11.4	mg/L	179	Standard
Fe	57	475.0	7.9	**0.5062**	0.132	26.1	mg/L	335	Standard
Sc-1	45	62358.1	1.0				mg/L	61029	Standard
Cl	35	34619.5	1.5				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	3197.0	3.6				ug/L	1547	Standard
P	31	18749.9	1.5				ug/L	15286	Standard
S	34	2010.1	3.0				ug/L	1572	Standard
Sr	88	305.0	20.5				ug/L	73	Standard
C	12	113.3	31.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	348.8	17.3				mg/L	6	Standard
Ho-1	165	175.0	2.9				mg/L	10	Standard
Er	166	163.3	52.8				mg/L	13	Standard
I	127	19115.4	2.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		146.759	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		107.762	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:21:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.720
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.008
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098603SDL WG574572-02
Sample Date/Time: Wednesday, June 29, 2016 13:22:37
Number of Replicates: 3
Autosampler Position: 318
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	107882.7	0.8				ug/L	74699	Standard
	Be	9	30.0	44.1	**0.0312**	0.010	32.7	ug/L	8	Standard
	Al	27	2675117.5	0.8	**12.5917**	0.213	1.7	ug/L	1493	Standard
	Sc	45	65760.7	1.1				ug/L	61029	Standard
	Ti	47	48.3	10.6	**0.0072**	0.012	160.2	ug/L	35	Standard
	V	51	2266.6	5.1	**0.0781**	0.015	19.5	ug/L	1435	Standard
	Cr	52	10143.4	2.6	**0.2243**	0.020	8.9	ug/L	7635	Standard
	Cr	53	413.3	17.1	**0.0775**	0.065	83.8	ug/L	363	Standard
	Mn	55	539662.1	1.4	**70.5366**	0.569	0.8	ug/L	747	Standard
	Co	59	1015.0	2.6	**0.1030**	0.006	5.9	ug/L	234	Standard
	Ni	60	769.4	2.6	**0.2509**	0.010	4.1	ug/L	215	Standard
	Cu	65	948.0	4.2	**0.3607**	0.016	4.4	ug/L	104	Standard
	Zn	66	2642.9	1.7	**1.7518**	0.018	1.1	ug/L	363	Standard
>	**Ge**	72	543000.4	1.8				ug/L	495142	Standard
	As	75	12.5	225.4	**0.0698**	0.023	32.5	ug/L	-50	Standard
	Se	82	15.8	49.7	**-0.0063**	0.066	1051.8	ug/L	14	Standard
	Se-1	77	45.3	17.1	**-0.0563**	0.096	170.3	ug/L	42	Standard
>	**Ga**	71	85.0	15.6				mg/L	42	Standard
	Rb	85	428.3	11.2				ug/L	17	Standard
	Y	89	460200.9	1.8				ug/L	417098	Standard
>	**Rh**	103	35.0	24.7				ug/L	13	Standard
	Mo	98	83.6	11.1	**0.0226**	0.003	14.3	ug/L	11	Standard
	Ag	107	112.7	14.2	**0.0027**	0.002	80.0	ug/L	68	Standard
	Cd	111	20.9	22.0	**0.0011**	0.002	176.9	mg/L	2	Standard
	Cd	114	53.5	39.5	**0.0073**	0.004	50.6	ug/L	12	Standard
>	**In**	115	649578.1	0.2				ug/L	587441	Standard
	Sn	118	553.3	9.5	**-0.0090**	0.008	92.4	ug/L	497	Standard
	Sb	123	558.9	31.9	**0.1041**	0.035	33.4	ug/L	78	Standard
	Ba	135	1064.4	0.7	**0.4543**	0.003	0.7	ug/L	24	Standard
	Ce	140	1928.5	4.3				ug/L	25	Standard
>	**Tb**	159	1049822.6	1.3				ug/L	939701	Standard
	Ho	165	40.0	33.1				ug/L	10	Standard
	Tl	203	40.7	35.8	**0.0026**	0.002	59.7	ug/L	4	Standard
	Tl	205	45.0	38.5	**0.0057**	0.002	36.9	ug/L	2	Standard
	Pb	206	372.0	6.6	**-0.0102**	0.004	38.7	ug/L	240	Standard
	Pb	207	294.7	6.9	**-0.0156**	0.004	25.6	ug/L	198	Standard
	Pb	208	1159.4	2.3	**-0.0199**	0.001	7.3	ug/L	883	Standard
	U	238	1206.0	0.9	**0.1490**	0.002	1.2	ug/L	3	Standard
>	**Bi**	209	598311.7	0.5				ug/L	523984	Standard

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:24:54
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.6400**	30.545	173.2	mg/L	0	Standard
Mg	24	363.3	29.0	**0.3711**	0.131	35.3	mg/L	37	Standard
K	39	63.3	27.7	**0.0157**	0.035	226.7	mg/L	38	Standard
Ca	43	110.0	12.0	**3.9202**	0.707	18.0	mg/L	10	Standard
Fe	54	190.5	10.6	**0.0117**	0.017	142.3	mg/L	179	Standard
Fe	57	268.3	12.0	**-0.2271**	0.110	48.6	mg/L	335	Standard
> **Sc-1**	45	65760.7	1.1				mg/L	61029	Standard
Cl	35	34122.4	1.0				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	2286.8	8.5				ug/L	1547	Standard
P	31	19529.2	1.3				ug/L	15286	Standard
S	34	2301.8	4.6				ug/L	1572	Standard
Sr	88	130.0	32.9				ug/L	73	Standard
C	12	83.3	6.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	60.8	35.4				mg/L	6	Standard
Ho-1	165	40.0	33.1				mg/L	10	Standard
Er	166	53.3	28.6				mg/L	13	Standard
I	127	5874.5	8.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		144.424	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.666	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098603SDL WG574572-02
Report Date/Time: Wednesday, June 29, 2016 13:24:54
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.578
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		114.185
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 13:25:50
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	96295.4	4.3				ug/L	74699	Standard
	Be	9	56739.5	4.7	**50.0144**	1.617	3.2	ug/L	8	Standard
	Al	27	8198811.7	1.9	**43.2979**	1.271	2.9	ug/L	1493	Standard
	Sc	45	60020.3	2.0				ug/L	61029	Standard
	Ti	47	38068.4	0.4	**97.0435**	0.523	0.5	ug/L	35	Standard
	V	51	407671.2	1.0	**52.0137**	0.145	0.3	ug/L	1435	Standard
	Cr	52	382711.8	0.9	**48.9253**	0.135	0.3	ug/L	7635	Standard
	Cr	53	47653.2	2.5	**49.5394**	1.027	2.1	ug/L	363	Standard
	Mn	55	366963.7	3.6	**50.8236**	1.362	2.7	ug/L	747	Standard
	Co	59	380449.1	2.9	**55.6813**	1.097	2.0	ug/L	234	Standard
	Ni	60	94585.7	1.1	**48.7401**	0.435	0.9	ug/L	215	Standard
	Cu	65	94763.8	1.5	**49.1432**	0.553	1.1	ug/L	104	Standard
	Zn	66	56752.8	2.3	**49.7949**	0.707	1.4	ug/L	363	Standard
>	Ge	72	511914.0	0.9				ug/L	495142	Standard
	As	75	57985.2	1.9	**50.1285**	0.568	1.1	ug/L	-50	Standard
	Se	82	5989.4	2.5	**52.6418**	0.964	1.8	ug/L	14	Standard
	Se-1	77	3921.2	2.5	**51.3010**	0.802	1.6	ug/L	42	Standard
>	Ga	71	220.0	19.4				mg/L	42	Standard
	Rb	85	765.0	2.0				ug/L	17	Standard
	Y	89	446724.4	3.3				ug/L	417098	Standard
>	Rh	103	20.0	43.3				ug/L	13	Standard
	Mo	98	280984.3	2.1	**102.0498**	2.098	2.1	ug/L	11	Standard
	Ag	107	335763.8	0.9	**49.8677**	1.010	2.0	ug/L	68	Standard
	Cd	111	105816.2	1.8	**48.9835**	0.569	1.2	mg/L	2	Standard
	Cd	114	257781.1	1.6	**48.6563**	0.589	1.2	ug/L	12	Standard
>	In	115	606960.1	2.8				ug/L	587441	Standard
	Sn	118	295693.2	2.0	**48.9061**	0.749	1.5	ug/L	497	Standard
	Sb	123	236523.1	0.9	**49.0438**	0.901	1.8	ug/L	78	Standard
	Ba	135	99992.2	1.5	**47.2781**	0.609	1.3	ug/L	24	Standard
	Ce	140	91.7	6.3				ug/L	25	Standard
>	Tb	159	985978.7	1.6				ug/L	939701	Standard
	Ho	165	11.7	89.2				ug/L	10	Standard
	Tl	203	433708.5	1.0	**49.3829**	0.257	0.5	ug/L	4	Standard
	Tl	205	391770.8	1.1	**51.3967**	0.163	0.3	ug/L	2	Standard
	Pb	206	284957.7	0.7	**50.0002**	0.477	1.0	ug/L	240	Standard
	Pb	207	255175.6	1.7	**50.3910**	0.211	0.4	ug/L	198	Standard
	Pb	208	1010912.5	1.9	**50.5251**	0.466	0.9	ug/L	883	Standard
	U	238	378758.8	0.6	**51.6451**	0.416	0.8	ug/L	3	Standard
>	Bi	209	558275.4	1.4				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 13:28:07
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4168.9	2.3	**5.4065**	0.021	0.4	mg/L	37	Standard
K	39	1983.5	2.5	**4.3432**	0.044	1.0	mg/L	38	Standard
Ca	43	106.7	26.7	**4.3258**	1.829	42.3	mg/L	10	Standard
Fe	54	6165.0	5.0	**5.1771**	0.265	5.1	mg/L	179	Standard
Fe	57	1720.1	9.9	**4.8644**	0.466	9.6	mg/L	335	Standard
> **Sc-1**	45	60020.3	2.0				mg/L	61029	Standard
Cl	35	28213.0	5.7				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	1853.5	23.0				ug/L	1547	Standard
P	31	20223.5	3.5				ug/L	15286	Standard
S	34	2575.2	6.1				ug/L	1572	Standard
Sr	88	78.3	26.6				ug/L	73	Standard
C	12	100.0	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	9.2	116.5				mg/L	6	Standard
Ho-1	165	11.7	89.2				mg/L	10	Standard
Er	166	16.7	91.7				mg/L	13	Standard
I	127	2436.9	25.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	100.029		
Al	27	86.596		
Sc	45			
Ti	47	97.044		
V	51	104.027		
Cr	52	97.851		
Cr	53			
Mn	55	101.647		
Co	59	111.363		
Ni	60	97.480		
Cu	65	98.286		
Zn	66	99.590		
> Ge	72		103.387	
As	75	100.257		
Se	82	105.284		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 13:28:07
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.050		
Ag	107	99.735		
Cd	111	97.967		
Cd	114			
In	115		103.323	
Sn	118	97.812		
Sb	123	98.088		
Ba	135	94.556		
Ce	140			
Tb	159			
Ho	165			
Tl	203	98.766		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.050		
U	238	103.290		
Bi	209		106.544	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 13:28:07
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 13:29:01
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94691.7	1.8				ug/L	74699	Standard
	Be	9	76.7	106.2	**0.0755**	0.071	94.2	ug/L	8	Standard
	Al	27	36843.1	136.1	**0.1814**	0.263	145.0	ug/L	1493	Standard
	Sc	45	60632.8	2.3				ug/L	61029	Standard
	Ti	47	43.0	36.1	**-0.0010**	0.038	3637.8	ug/L	35	Standard
	V	51	1519.7	11.1	**-0.0035**	0.019	532.9	ug/L	1435	Standard
	Cr	52	7045.6	5.2	**-0.1157**	0.044	37.8	ug/L	7635	Standard
	Cr	53	391.7	11.6	**0.0747**	0.044	59.2	ug/L	363	Standard
	Mn	55	7226.9	125.7	**0.8342**	1.228	147.2	ug/L	747	Standard
	Co	59	517.7	90.4	**0.0371**	0.066	179.1	ug/L	234	Standard
	Ni	60	246.3	39.3	**0.0014**	0.048	3317.1	ug/L	215	Standard
	Cu	65	191.7	56.6	**-0.0057**	0.054	949.5	ug/L	104	Standard
	Zn	66	417.3	10.8	**-0.0898**	0.036	40.1	ug/L	363	Standard
>	**Ge**	72	518398.0	1.4				ug/L	495142	Standard
	As	75	-30.5	76.0	**0.0334**	0.020	59.9	ug/L	-50	Standard
	Se	82	13.4	12.4	**-0.0212**	0.016	76.0	ug/L	14	Standard
	Se-1	77	53.3	15.9	**0.0746**	0.106	141.7	ug/L	42	Standard
>	**Ga**	71	155.0	18.0				mg/L	42	Standard
	Rb	85	21.7	93.3				ug/L	17	Standard
	Y	89	436745.4	0.7				ug/L	417098	Standard
>	**Rh**	103	15.0	33.3				ug/L	13	Standard
	Mo	98	315.2	32.1	**0.1069**	0.036	33.6	ug/L	11	Standard
	Ag	107	222.7	75.6	**0.0196**	0.025	125.0	ug/L	68	Standard
	Cd	111	57.8	114.5	**0.0184**	0.030	163.6	mg/L	2	Standard
	Cd	114	159.0	130.1	**0.0274**	0.038	140.1	ug/L	12	Standard
>	**In**	115	616374.3	0.2				ug/L	587441	Standard
	Sn	118	845.0	17.1	**0.0432**	0.024	55.2	ug/L	497	Standard
	Sb	123	593.6	13.2	**0.1170**	0.016	13.9	ug/L	78	Standard
	Ba	135	102.7	102.4	**0.0317**	0.049	153.8	ug/L	24	Standard
	Ce	140	58.3	55.8				ug/L	25	Standard
>	**Tb**	159	995160.6	1.6				ug/L	939701	Standard
	Ho	165	15.0	100.0				ug/L	10	Standard
	Tl	203	219.0	114.2	**0.0226**	0.028	122.3	ug/L	4	Standard
	Tl	205	381.7	141.6	**0.0491**	0.069	140.6	ug/L	2	Standard
	Pb	206	388.0	38.6	**-0.0042**	0.025	597.1	ug/L	240	Standard
	Pb	207	336.0	35.9	**-0.0047**	0.023	485.4	ug/L	198	Standard
	Pb	208	2012.1	79.6	**0.0246**	0.078	315.7	ug/L	883	Standard
	U	238	395.3	145.1	**0.0482**	0.076	157.9	ug/L	3	Standard
>	**Bi**	209	567355.9	1.2				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 13:31:18
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	58.3	34.6	**0.0111**	0.028	248.5	mg/L	37	Standard
K	39	40.0	12.5	**-0.0251**	0.012	46.0	mg/L	38	Standard
Ca	43	31.7	9.1	**-0.5078**	0.213	41.9	mg/L	10	Standard
Fe	54	155.5	6.6	**-0.0056**	0.009	155.5	mg/L	179	Standard
Fe	57	230.0	7.8	**-0.2862**	0.078	27.3	mg/L	335	Standard
> **Sc-1**	45	60632.8	2.3				mg/L	61029	Standard
Cl	35	27178.4	0.6				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	1736.8	2.2				ug/L	1547	Standard
P	31	19502.5	1.8				ug/L	15286	Standard
S	34	2301.8	3.7				ug/L	1572	Standard
Sr	88	75.0	26.7				ug/L	73	Standard
C	12	86.7	17.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	12.7	93.2				mg/L	6	Standard
Ho-1	165	15.0	100.0				mg/L	10	Standard
Er	166	13.3	86.6				mg/L	13	Standard
I	127	1935.1	25.0				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		104.697	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 13:31:18
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.925
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.277
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Mn	55	

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 13:31:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098605
Sample Date/Time: Wednesday, June 29, 2016 13:32:14
Number of Replicates: 3
Autosampler Position: 319
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93251.1	1.7				ug/L	74699	Standard
	Be	9	326.7	10.4	**0.3052**	0.035	11.6	ug/L	8	Standard
	Al	27	19740200.2	0.7	**107.5996**	1.185	1.1	ug/L	1493	Standard
	Sc	45	60365.1	2.3				ug/L	61029	Standard
	Ti	47	450.3	3.0	**1.0603**	0.030	2.8	ug/L	35	Standard
	V	51	4644.9	8.0	**0.4097**	0.047	11.5	ug/L	1435	Standard
	Cr	52	12104.2	0.3	**0.5838**	0.010	1.7	ug/L	7635	Standard
	Cr	53	928.4	3.1	**0.6586**	0.026	3.9	ug/L	363	Standard
	Mn	55	98559.3	4.8	**13.7841**	0.650	4.7	ug/L	747	Standard
	Co	59	3296.0	4.6	**0.4539**	0.021	4.7	ug/L	234	Standard
	Ni	60	3163.0	1.8	**1.5389**	0.038	2.4	ug/L	215	Standard
	Cu	65	32310.7	0.8	**16.9827**	0.169	1.0	ug/L	104	Standard
	Zn	66	47105.4	0.5	**41.9893**	0.365	0.9	ug/L	363	Standard
>	**Ge**	72	503086.3	0.5				ug/L	495142	Standard
	As	75	294.8	15.7	**0.3186**	0.040	12.5	ug/L	-50	Standard
	Se	82	52.3	21.7	**0.3309**	0.101	30.6	ug/L	14	Standard
	Se-1	77	70.7	11.4	**0.3303**	0.114	34.4	ug/L	42	Standard
>	**Ga**	71	295.0	9.0				mg/L	42	Standard
	Rb	85	9109.4	3.7				ug/L	17	Standard
	Y	89	448266.6	2.2				ug/L	417098	Standard
>	**Rh**	103	83.3	28.4				ug/L	13	Standard
	Mo	98	1843.0	5.2	**0.6500**	0.034	5.3	ug/L	11	Standard
	Ag	107	226.3	74.5	**0.0200**	0.024	121.9	ug/L	68	Standard
	Cd	111	104.1	84.7	**0.0392**	0.040	101.3	mg/L	2	Standard
	Cd	114	254.7	74.3	**0.0449**	0.035	77.3	ug/L	12	Standard
>	**In**	115	619535.9	1.6				ug/L	587441	Standard
	Sn	118	1750.1	13.2	**0.1895**	0.038	20.1	ug/L	497	Standard
	Sb	123	790.9	17.9	**0.1564**	0.028	18.0	ug/L	78	Standard
	Ba	135	72683.9	1.3	**33.6664**	0.969	2.9	ug/L	24	Standard
	Ce	140	5139.2	3.4				ug/L	25	Standard
>	**Tb**	159	985815.7	2.2				ug/L	939701	Standard
	Ho	165	223.3	9.3				ug/L	10	Standard
	Tl	203	317.0	96.7	**0.0353**	0.035	100.6	ug/L	4	Standard
	Tl	205	193.3	47.9	**0.0263**	0.012	46.4	ug/L	2	Standard
	Pb	206	1382.4	8.3	**0.1795**	0.018	10.0	ug/L	240	Standard
	Pb	207	1152.7	10.0	**0.1653**	0.022	13.1	ug/L	198	Standard
	Pb	208	4722.3	7.7	**0.1699**	0.018	10.3	ug/L	883	Standard
	U	238	9861.9	3.5	**1.3841**	0.043	3.1	ug/L	3	Standard
>	**Bi**	209	540670.7	1.7				ug/L	523984	Standard

Sample ID: L1606098605
Report Date/Time: Wednesday, June 29, 2016 13:34:31
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	6986.6	1.5	**9.0573**	0.310	3.4	mg/L	37	Standard
K	39	308.3	17.9	**0.5733**	0.109	19.0	mg/L	38	Standard
Ca	43	606.7	4.1	**36.2577**	2.485	6.9	mg/L	10	Standard
Fe	54	348.6	15.4	**0.1605**	0.045	28.3	mg/L	179	Standard
Fe	57	585.0	8.4	**0.9378**	0.194	20.7	mg/L	335	Standard
> **Sc-1**	45	60365.1	2.3				mg/L	61029	Standard
Cl	35	31561.5	3.4				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	4964.2	11.3				ug/L	1547	Standard
P	31	20053.3	1.8				ug/L	15286	Standard
S	34	2155.2	4.8				ug/L	1572	Standard
Sr	88	343.3	18.3				ug/L	73	Standard
C	12	163.3	34.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	294.3	16.7				mg/L	6	Standard
Ho-1	165	223.3	9.3				mg/L	10	Standard
Er	166	120.0	16.7				mg/L	13	Standard
I	127	9371.2	3.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		124.836	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.604	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098605
Report Date/Time: Wednesday, June 29, 2016 13:34:31
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	105.463
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	103.184
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098607
Sample Date/Time: Wednesday, June 29, 2016 13:35:26
Number of Replicates: 3
Autosampler Position: 320
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96566.8	1.9				ug/L	74699	Standard
	Be	9	85.0	29.4	**0.0821**	0.021	25.7	ug/L	8	Standard
	Al	27	20333697.9	3.0	**107.0230**	2.931	2.7	ug/L	1493	Standard
	Sc	45	62230.9	2.2				ug/L	61029	Standard
	Ti	47	287.7	7.2	**0.6119**	0.056	9.2	ug/L	35	Standard
	V	51	3941.3	2.6	**0.3001**	0.017	5.7	ug/L	1435	Standard
	Cr	52	11512.0	2.6	**0.4518**	0.053	11.8	ug/L	7635	Standard
	Cr	53	838.4	4.6	**0.5317**	0.044	8.4	ug/L	363	Standard
	Mn	55	502202.8	0.8	**68.3542**	0.721	1.1	ug/L	747	Standard
	Co	59	4570.7	1.4	**0.6201**	0.010	1.6	ug/L	234	Standard
	Ni	60	3293.0	0.7	**1.5464**	0.014	0.9	ug/L	215	Standard
	Cu	65	6694.1	1.0	**3.3123**	0.070	2.1	ug/L	104	Standard
	Zn	66	11579.4	0.2	**9.6129**	0.092	1.0	ug/L	363	Standard
>	**Ge**	72	521399.1	1.0				ug/L	495142	Standard
	As	75	65.2	31.2	**0.1148**	0.017	14.9	ug/L	-50	Standard
	Se	82	28.9	16.0	**0.1117**	0.041	36.5	ug/L	14	Standard
	Se-1	77	63.7	8.1	**0.2056**	0.068	33.2	ug/L	42	Standard
>	**Ga**	71	136.7	35.5				mg/L	42	Standard
	Rb	85	5054.2	2.4				ug/L	17	Standard
	Y	89	452894.0	2.8				ug/L	417098	Standard
>	**Rh**	103	165.0	28.9				ug/L	13	Standard
	Mo	98	1211.0	1.7	**0.4303**	0.008	1.9	ug/L	11	Standard
	Ag	107	106.3	11.0	**0.0028**	0.002	64.6	ug/L	68	Standard
	Cd	111	72.3	5.5	**0.0253**	0.002	7.2	mg/L	2	Standard
	Cd	114	189.4	18.5	**0.0333**	0.006	19.3	ug/L	12	Standard
>	**In**	115	612132.4	0.4				ug/L	587441	Standard
	Sn	118	1080.0	6.7	**0.0827**	0.012	15.0	ug/L	497	Standard
	Sb	123	567.1	13.2	**0.1124**	0.015	13.5	ug/L	78	Standard
	Ba	135	35457.1	0.1	**16.6089**	0.050	0.3	ug/L	24	Standard
	Ce	140	1880.1	4.7				ug/L	25	Standard
>	**Tb**	159	996852.4	0.7				ug/L	939701	Standard
	Ho	165	58.3	17.8				ug/L	10	Standard
	Tl	203	60.0	16.9	**0.0051**	0.001	23.4	ug/L	4	Standard
	Tl	205	60.0	16.7	**0.0081**	0.001	15.8	ug/L	2	Standard
	Pb	206	636.0	0.7	**0.0415**	0.001	3.4	ug/L	240	Standard
	Pb	207	529.0	3.2	**0.0355**	0.003	9.7	ug/L	198	Standard
	Pb	208	2291.4	2.0	**0.0416**	0.003	7.7	ug/L	883	Standard
	U	238	20180.8	2.1	**2.7702**	0.079	2.9	ug/L	3	Standard
>	**Bi**	209	553749.9	0.8				ug/L	523984	Standard

Sample ID: L1606098607
Report Date/Time: Wednesday, June 29, 2016 13:37:42
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**57.0696**	57.922	101.5	mg/L	0	Standard
Mg	24	5551.0	3.6	**6.9630**	0.244	3.5	mg/L	37	Standard
K	39	316.7	11.2	**0.5727**	0.081	14.1	mg/L	38	Standard
Ca	43	636.7	7.9	**36.9594**	3.698	10.0	mg/L	10	Standard
Fe	54	176.9	14.5	**0.0085**	0.018	212.8	mg/L	179	Standard
Fe	57	531.7	3.6	**0.6989**	0.084	12.0	mg/L	335	Standard
> **Sc-1**	45	62230.9	2.2				mg/L	61029	Standard
Cl	35	31386.5	1.7				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	3093.6	2.6				ug/L	1547	Standard
P	31	20912.8	5.9				ug/L	15286	Standard
S	34	2228.5	4.5				ug/L	1572	Standard
Sr	88	703.3	5.1				ug/L	73	Standard
C	12	126.7	40.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	102.2	30.1				mg/L	6	Standard
Ho-1	165	58.3	17.8				mg/L	10	Standard
Er	166	93.3	6.2				mg/L	13	Standard
I	127	29892.9	5.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		129.275	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		105.303	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098607
Report Date/Time: Wednesday, June 29, 2016 13:37:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.203
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.681
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098609
Sample Date/Time: Wednesday, June 29, 2016 13:38:37
Number of Replicates: 3
Autosampler Position: 321
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96550.0	0.2				ug/L	74699	Standard
	Be	9	53.3	14.3	**0.0545**	0.007	12.5	ug/L	8	Standard
	Al	27	22878376.4	1.6	**120.4307**	1.945	1.6	ug/L	1493	Standard
	Sc	45	62031.7	0.8				ug/L	61029	Standard
	Ti	47	1850.4	1.6	**4.6226**	0.192	4.2	ug/L	35	Standard
	V	51	5461.8	0.6	**0.5052**	0.022	4.3	ug/L	1435	Standard
	Cr	52	13396.9	1.2	**0.7280**	0.058	8.0	ug/L	7635	Standard
	Cr	53	1881.8	4.3	**1.6444**	0.138	8.4	ug/L	363	Standard
	Mn	55	79733.3	1.7	**10.9451**	0.135	1.2	ug/L	747	Standard
	Co	59	2496.2	2.4	**0.3289**	0.007	2.0	ug/L	234	Standard
	Ni	60	4821.8	0.8	**2.3714**	0.068	2.9	ug/L	215	Standard
	Cu	65	4208.3	2.7	**2.0873**	0.088	4.2	ug/L	104	Standard
	Zn	66	20900.1	0.5	**18.0858**	0.567	3.1	ug/L	363	Standard
>	**Ge**	72	511251.4	2.6				ug/L	495142	Standard
	As	75	201.6	14.3	**0.2338**	0.023	9.7	ug/L	-50	Standard
	Se	82	54.1	5.4	**0.3394**	0.016	4.7	ug/L	14	Standard
	Se-1	77	103.0	11.2	**0.7418**	0.124	16.8	ug/L	42	Standard
>	**Ga**	71	510.0	8.0				mg/L	42	Standard
	Rb	85	13392.6	2.9				ug/L	17	Standard
	Y	89	437578.5	1.9				ug/L	417098	Standard
>	**Rh**	103	115.0	21.7				ug/L	13	Standard
	Mo	98	379.6	2.6	**0.1321**	0.004	2.9	ug/L	11	Standard
	Ag	107	107.7	0.5	**0.0031**	0.000	8.0	ug/L	68	Standard
	Cd	111	160.7	1.3	**0.0664**	0.001	1.0	mg/L	2	Standard
	Cd	114	395.9	2.3	**0.0726**	0.002	3.0	ug/L	12	Standard
>	**In**	115	607119.7	1.0				ug/L	587441	Standard
	Sn	118	1270.1	2.5	**0.1157**	0.007	5.8	ug/L	497	Standard
	Sb	123	941.4	6.3	**0.1909**	0.012	6.2	ug/L	78	Standard
	Ba	135	18081.1	0.3	**8.5321**	0.064	0.8	ug/L	24	Standard
	Ce	140	9564.7	2.3				ug/L	25	Standard
>	**Tb**	159	994932.9	1.3				ug/L	939701	Standard
	Ho	165	368.3	5.1				ug/L	10	Standard
	Tl	203	147.3	4.5	**0.0153**	0.001	4.3	ug/L	4	Standard
	Tl	205	150.0	15.3	**0.0202**	0.003	16.2	ug/L	2	Standard
	Pb	206	1807.1	1.2	**0.2512**	0.001	0.3	ug/L	240	Standard
	Pb	207	1489.4	1.7	**0.2291**	0.005	2.2	ug/L	198	Standard
	Pb	208	6035.1	2.4	**0.2326**	0.004	1.7	ug/L	883	Standard
	U	238	2872.6	2.3	**0.3933**	0.005	1.2	ug/L	3	Standard
>	**Bi**	209	549776.1	1.1				ug/L	523984	Standard

Sample ID: L1606098609
Report Date/Time: Wednesday, June 29, 2016 13:40:55
Page 1

Approved: July 04, 2016

Na	23	6.7	114.6	**75.5001**	86.744	114.9	mg/L	0	Standard
Mg	24	5847.8	3.3	**7.3629**	0.301	4.1	mg/L	37	Standard
K	39	155.0	19.6	**0.2229**	0.066	29.5	mg/L	38	Standard
Ca	43	803.4	3.4	**47.4095**	2.024	4.3	mg/L	10	Standard
Fe	54	459.0	5.6	**0.2447**	0.024	9.9	mg/L	179	Standard
Fe	57	691.7	11.1	**1.2400**	0.273	22.0	mg/L	335	Standard
> **Sc-1**	45	62031.7	0.8				mg/L	61029	Standard
Cl	35	35055.2	1.5				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	8335.6	0.3				ug/L	1547	Standard
P	31	19207.2	4.0				ug/L	15286	Standard
S	34	2091.8	7.3				ug/L	1572	Standard
Sr	88	440.0	15.9				ug/L	73	Standard
C	12	146.7	7.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	580.9	19.4				mg/L	6	Standard
Ho-1	165	368.3	5.1				mg/L	10	Standard
Er	166	330.0	13.9				mg/L	13	Standard
I	127	13649.5	1.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		129.252	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.253	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098609
Report Date/Time: Wednesday, June 29, 2016 13:40:55
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	103.350
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.922
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606098609
Report Date/Time: Wednesday, June 29, 2016 13:40:55
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098611
Sample Date/Time: Wednesday, June 29, 2016 13:41:50
Number of Replicates: 3
Autosampler Position: 322
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94928.2	1.7				ug/L	74699	Standard
	Be	9	611.7	2.1	**0.5543**	0.002	0.4	ug/L	8	Standard
	Al	27	35531075.2	1.8	**190.3104**	6.477	3.4	ug/L	1493	Standard
	Sc	45	58926.0	1.1				ug/L	61029	Standard
	Ti	47	1221.7	3.3	**3.0767**	0.071	2.3	ug/L	35	Standard
	V	51	7317.6	1.4	**0.7622**	0.017	2.2	ug/L	1435	Standard
	Cr	52	16086.9	1.7	**1.1219**	0.012	1.1	ug/L	7635	Standard
	Cr	53	9996.6	2.0	**10.3612**	0.096	0.9	ug/L	363	Standard
	Mn	55	583301.8	0.5	**82.6724**	0.812	1.0	ug/L	747	Standard
	Co	59	14960.7	0.8	**2.2024**	0.031	1.4	ug/L	234	Standard
	Ni	60	6882.6	4.3	**3.5089**	0.114	3.2	ug/L	215	Standard
	Cu	65	67223.1	1.5	**35.5977**	0.129	0.4	ug/L	104	Standard
	Zn	66	37202.6	1.0	**33.2122**	0.242	0.7	ug/L	363	Standard
>	**Ge**	72	500908.0	1.1				ug/L	495142	Standard
	As	75	1307.8	5.6	**1.2134**	0.054	4.4	ug/L	-50	Standard
	Se	82	362.7	2.0	**3.1282**	0.043	1.4	ug/L	14	Standard
	Se-1	77	555.0	5.1	**6.8873**	0.307	4.5	ug/L	42	Standard
>	**Ga**	71	435.0	11.9				mg/L	42	Standard
	Rb	85	18344.4	2.4				ug/L	17	Standard
	Y	89	436166.7	1.7				ug/L	417098	Standard
>	**Rh**	103	445.0	7.4				ug/L	13	Standard
	Mo	98	3054.9	0.8	**1.1522**	0.023	2.0	ug/L	11	Standard
	Ag	107	121.7	12.3	**0.0060**	0.002	41.3	ug/L	68	Standard
	Cd	111	41.7	14.4	**0.0123**	0.003	24.8	mg/L	2	Standard
	Cd	114	142.6	14.6	**0.0260**	0.004	14.4	ug/L	12	Standard
>	**In**	115	581588.5	1.2				ug/L	587441	Standard
	Sn	118	2201.8	2.3	**0.2861**	0.011	3.9	ug/L	497	Standard
	Sb	123	1347.5	5.7	**0.2874**	0.019	6.6	ug/L	78	Standard
	Ba	135	63997.2	1.5	**31.5730**	0.849	2.7	ug/L	24	Standard
	Ce	140	22939.0	3.1				ug/L	25	Standard
>	**Tb**	159	946929.2	0.6				ug/L	939701	Standard
	Ho	165	348.3	12.2				ug/L	10	Standard
	Tl	203	270.7	7.5	**0.0312**	0.002	6.9	ug/L	4	Standard
	Tl	205	231.7	2.5	**0.0327**	0.001	2.0	ug/L	2	Standard
	Pb	206	3400.0	1.8	**0.5672**	0.008	1.4	ug/L	240	Standard
	Pb	207	2794.6	0.7	**0.5207**	0.008	1.5	ug/L	198	Standard
	Pb	208	11397.0	1.2	**0.5356**	0.008	1.6	ug/L	883	Standard
	U	238	7265.4	1.1	**1.0543**	0.007	0.7	ug/L	3	Standard
>	**Bi**	209	522379.2	0.9				ug/L	523984	Standard

Sample ID: L1606098611
Report Date/Time: Wednesday, June 29, 2016 13:44:07
Page 1

Approved: July 04, 2016

Na	23	10.0	50.0	**119.8009**	60.997	50.9	mg/L	0	Standard
Mg	24	10421.9	2.0	**13.8711**	0.425	3.1	mg/L	37	Standard
K	39	453.3	5.1	**0.9237**	0.053	5.7	mg/L	38	Standard
Ca	43	1046.7	6.1	**65.9862**	4.937	7.5	mg/L	10	Standard
Fe	54	728.6	2.4	**0.5015**	0.016	3.2	mg/L	179	Standard
Fe	57	860.0	12.2	**1.9559**	0.399	20.4	mg/L	335	Standard
> **Sc-1**	45	58926.0	1.1				mg/L	61029	Standard
Cl	35	32129.4	1.2				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	75514.4	2.1				ug/L	1547	Standard
P	31	20090.0	4.4				ug/L	15286	Standard
S	34	2098.5	7.9				ug/L	1572	Standard
Sr	88	1583.4	4.7				ug/L	73	Standard
C	12	130.0	7.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	561.0	7.3				mg/L	6	Standard
Ho-1	165	348.3	12.2				mg/L	10	Standard
Er	166	256.7	6.0				mg/L	13	Standard
I	127	84570.2	9.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		127.081	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		101.164	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098611
Report Date/Time: Wednesday, June 29, 2016 13:44:07
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.004
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.694
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098613
Sample Date/Time: Wednesday, June 29, 2016 13:45:02
Number of Replicates: 3
Autosampler Position: 323
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101811.1	2.5				ug/L	74699	Standard
	Be	9	780.0	6.8	**0.6573**	0.030	4.5	ug/L	8	Standard
	Al	27	69460342.3	2.9	**346.7492**	1.873	0.5	ug/L	1493	Standard
	Sc	45	66852.2	1.8				ug/L	61029	Standard
	Ti	47	16395.2	2.4	**40.4139**	0.750	1.9	ug/L	35	Standard
	V	51	44929.4	0.9	**5.3782**	0.102	1.9	ug/L	1435	Standard
	Cr	52	38038.0	1.4	**3.7847**	0.045	1.2	ug/L	7635	Standard
	Cr	53	5357.6	0.6	**5.1013**	0.082	1.6	ug/L	363	Standard
	Mn	55	1205027.8	2.1	**162.0424**	5.610	3.5	ug/L	747	Standard
	Co	59	15267.0	2.9	**2.1291**	0.080	3.8	ug/L	234	Standard
	Ni	60	13379.6	2.6	**6.5697**	0.065	1.0	ug/L	215	Standard
	Cu	65	74766.0	1.3	**37.5331**	0.708	1.9	ug/L	104	Standard
	Zn	66	182914.0	1.7	**156.4214**	2.327	1.5	ug/L	363	Standard
>	**Ge**	72	528583.9	2.0				ug/L	495142	Standard
	As	75	2539.4	2.2	**2.1832**	0.013	0.6	ug/L	-50	Standard
	Se	82	108.7	5.6	**0.7905**	0.071	9.0	ug/L	14	Standard
	Se-1	77	150.3	9.2	**1.3050**	0.161	12.3	ug/L	42	Standard
>	**Ga**	71	5017.5	0.9				mg/L	42	Standard
	Rb	85	95755.1	2.6				ug/L	17	Standard
	Y	89	484844.4	1.2				ug/L	417098	Standard
>	**Rh**	103	451.7	17.2				ug/L	13	Standard
	Mo	98	2711.0	1.7	**0.9675**	0.010	1.1	ug/L	11	Standard
	Ag	107	547.0	8.0	**0.0673**	0.005	7.7	ug/L	68	Standard
	Cd	111	100.9	10.4	**0.0383**	0.005	14.0	mg/L	2	Standard
	Cd	114	263.2	6.1	**0.0470**	0.003	5.8	ug/L	12	Standard
>	**In**	115	613966.8	2.2				ug/L	587441	Standard
	Sn	118	1991.8	3.6	**0.2317**	0.015	6.6	ug/L	497	Standard
	Sb	123	971.4	9.0	**0.1947**	0.014	7.1	ug/L	78	Standard
	Ba	135	97214.2	0.8	**45.4386**	0.665	1.5	ug/L	24	Standard
	Ce	140	163522.7	0.6				ug/L	25	Standard
>	**Tb**	159	1011575.1	2.5				ug/L	939701	Standard
	Ho	165	4682.4	4.2				ug/L	10	Standard
	Tl	203	573.7	3.0	**0.0642**	0.003	4.6	ug/L	4	Standard
	Tl	205	525.0	7.2	**0.0698**	0.005	7.8	ug/L	2	Standard
	Pb	206	15938.7	1.8	**2.7562**	0.018	0.6	ug/L	240	Standard
	Pb	207	13053.6	0.6	**2.5367**	0.043	1.7	ug/L	198	Standard
	Pb	208	53401.0	1.9	**2.6245**	0.013	0.5	ug/L	883	Standard
	U	238	38818.0	2.5	**5.3391**	0.057	1.1	ug/L	3	Standard
>	**Bi**	209	553002.2	2.3				ug/L	523984	Standard

Sample ID: L1606098613
Report Date/Time: Wednesday, June 29, 2016 13:47:19
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**88.0394**	31.556	35.8	mg/L	0	Standard
Mg	24	10193.4	4.3	**11.9448**	0.378	3.2	mg/L	37	Standard
K	39	798.4	2.0	**1.4974**	0.060	4.0	mg/L	38	Standard
Ca	43	1498.4	4.4	**83.9112**	4.508	5.4	mg/L	10	Standard
Fe	54	4251.3	3.0	**3.1535**	0.141	4.5	mg/L	179	Standard
Fe	57	1956.8	0.6	**4.9983**	0.081	1.6	mg/L	335	Standard
> **Sc-1**	45	66852.2	1.8				mg/L	61029	Standard
Cl	35	38555.7	1.3				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	16405.6	4.1				ug/L	1547	Standard
P	31	22243.0	1.3				ug/L	15286	Standard
S	34	1961.8	3.3				ug/L	1572	Standard
Sr	88	2096.8	8.9				ug/L	73	Standard
C	12	170.0	21.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	107.9				mg/L	3	Standard
Dy	164	7687.7	5.1				mg/L	6	Standard
Ho-1	165	4682.4	4.2				mg/L	10	Standard
Er	166	4197.2	1.1				mg/L	13	Standard
I	127	34873.4	0.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		136.295	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		106.754	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098613
Report Date/Time: Wednesday, June 29, 2016 13:47:19
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.515
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.538
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098613
Report Date/Time: Wednesday, June 29, 2016 13:47:19
Page 3

Approved: July 04, 2016

Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098615
Sample Date/Time: Wednesday, June 29, 2016 13:48:13
Number of Replicates: 3
Autosampler Position: 324
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	104439.3	4.8				ug/L	74699	Standard
	Be	9	536.7	12.4	**0.4427**	0.037	8.4	ug/L	8	Standard
	Al	27	77862989.2	8.3	**378.5121**	14.291	3.8	ug/L	1493	Standard
	Sc	45	58982.2	11.3				ug/L	61029	Standard
	Ti	47	2461.2	11.7	**6.5990**	0.434	6.6	ug/L	35	Standard
	V	51	4945.2	8.0	**0.4817**	0.017	3.5	ug/L	1435	Standard
	Cr	52	13134.4	7.3	**0.8107**	0.065	8.0	ug/L	7635	Standard
	Cr	53	1810.1	8.3	**1.6977**	0.117	6.9	ug/L	363	Standard
	Mn	55	1145096.1	8.2	**169.9719**	4.564	2.7	ug/L	747	Standard
	Co	59	4019.2	11.3	**0.5913**	0.032	5.4	ug/L	234	Standard
	Ni	60	7177.1	13.9	**3.8310**	0.314	8.2	ug/L	215	Standard
	Cu	65	41403.8	8.7	**22.8999**	0.588	2.6	ug/L	104	Standard
	Zn	66	43831.2	7.7	**41.0562**	0.750	1.8	ug/L	363	Standard
>	**Ge**	72	478398.7	6.6				ug/L	495142	Standard
	As	75	638.2	9.5	**0.6495**	0.045	6.9	ug/L	-50	Standard
	Se	82	102.9	17.3	**0.8353**	0.172	20.6	ug/L	14	Standard
	Se-1	77	94.0	19.4	**0.7027**	0.185	26.4	ug/L	42	Standard
>	**Ga**	71	463.3	11.0				mg/L	42	Standard
	Rb	85	23055.9	7.7				ug/L	17	Standard
	Y	89	432649.5	7.6				ug/L	417098	Standard
>	**Rh**	103	1188.4	10.3				ug/L	13	Standard
	Mo	98	1350.9	6.9	**0.5044**	0.013	2.5	ug/L	11	Standard
	Ag	107	139.0	6.4	**0.0086**	0.002	19.3	ug/L	68	Standard
	Cd	111	44.6	6.8	**0.0137**	0.002	16.4	mg/L	2	Standard
	Cd	114	129.2	31.8	**0.0237**	0.010	43.4	ug/L	12	Standard
>	**In**	115	583863.5	7.1				ug/L	587441	Standard
	Sn	118	2513.5	5.1	**0.3387**	0.012	3.5	ug/L	497	Standard
	Sb	123	705.1	6.8	**0.1478**	0.002	1.6	ug/L	78	Standard
	Ba	135	26452.2	7.2	**12.9866**	0.138	1.1	ug/L	24	Standard
	Ce	140	27398.2	9.1				ug/L	25	Standard
>	**Tb**	159	958605.6	6.1				ug/L	939701	Standard
	Ho	165	645.0	4.0				ug/L	10	Standard
	Tl	203	134.0	9.5	**0.0146**	0.001	7.1	ug/L	4	Standard
	Tl	205	118.3	6.5	**0.0169**	0.002	13.2	ug/L	2	Standard
	Pb	206	2508.9	7.1	**0.4009**	0.002	0.4	ug/L	240	Standard
	Pb	207	1976.5	6.7	**0.3487**	0.007	2.0	ug/L	198	Standard
	Pb	208	8160.9	6.7	**0.3634**	0.004	1.2	ug/L	883	Standard
	U	238	17292.6	7.2	**2.5205**	0.006	0.2	ug/L	3	Standard
>	**Bi**	209	521380.1	7.3				ug/L	523984	Standard

Sample ID: L1606098615
Report Date/Time: Wednesday, June 29, 2016 13:50:30
Page 1

Approved: July 04, 2016

Na	23	15.0	57.7	**176.9361**	92.757	52.4	mg/L	0	Standard
Mg	24	28440.1	8.6	**38.0068**	1.323	3.5	mg/L	37	Standard
K	39	673.3	1.9	**1.4418**	0.213	14.8	mg/L	38	Standard
Ca	43	1596.8	5.0	**102.4296**	8.312	8.1	mg/L	10	Standard
Fe	54	1077.2	5.0	**0.8118**	0.070	8.7	mg/L	179	Standard
Fe	57	1203.4	10.1	**3.1714**	0.327	10.3	mg/L	335	Standard
> **Sc-1**	45	58982.2	11.3				mg/L	61029	Standard
Cl	35	41482.2	7.1				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	15648.1	7.3				ug/L	1547	Standard
P	31	20014.9	6.8				ug/L	15286	Standard
S	34	1913.5	9.5				ug/L	1572	Standard
Sr	88	5429.3	11.3				ug/L	73	Standard
C	12	203.3	27.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	1127.2	20.3				mg/L	6	Standard
Ho-1	165	645.0	4.0				mg/L	10	Standard
Er	166	546.7	6.4				mg/L	13	Standard
I	127	112969.0	6.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		139.814	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.618	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098615
Report Date/Time: Wednesday, June 29, 2016 13:50:30
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		99.391
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.503
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098615
Report Date/Time: Wednesday, June 29, 2016 13:50:30
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098617
Sample Date/Time: Wednesday, June 29, 2016 13:51:25
Number of Replicates: 3
Autosampler Position: 325
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95101.0	2.9				ug/L	74699	Standard
	Be	9	61.7	9.4	**0.0626**	0.004	5.8	ug/L	8	Standard
	Al	27	16761639.5	0.8	**89.6118**	1.975	2.2	ug/L	1493	Standard
	Sc	45	53828.8	0.6				ug/L	61029	Standard
	Ti	47	756.4	12.3	**2.0144**	0.253	12.6	ug/L	35	Standard
	V	51	3040.4	3.6	**0.2328**	0.013	5.5	ug/L	1435	Standard
	Cr	52	11892.0	0.9	**0.6842**	0.022	3.3	ug/L	7635	Standard
	Cr	53	1066.7	2.7	**0.8986**	0.036	4.1	ug/L	363	Standard
	Mn	55	57447.7	1.9	**8.6332**	0.089	1.0	ug/L	747	Standard
	Co	59	1276.7	2.2	**0.1685**	0.005	2.7	ug/L	234	Standard
	Ni	60	2114.5	5.4	**1.0778**	0.056	5.2	ug/L	215	Standard
	Cu	65	5578.0	1.5	**3.0859**	0.022	0.7	ug/L	104	Standard
	Zn	66	8715.8	0.2	**8.0380**	0.061	0.8	ug/L	363	Standard
>	**Ge**	72	465226.7	0.9				ug/L	495142	Standard
	As	75	386.7	5.7	**0.4269**	0.018	4.2	ug/L	-50	Standard
	Se	82	42.7	8.0	**0.2761**	0.036	13.2	ug/L	14	Standard
	Se-1	77	67.0	16.1	**0.3532**	0.150	42.5	ug/L	42	Standard
>	**Ga**	71	183.3	18.2				mg/L	42	Standard
	Rb	85	8882.6	2.7				ug/L	17	Standard
	Y	89	409246.0	2.8				ug/L	417098	Standard
>	**Rh**	103	295.0	13.6				ug/L	13	Standard
	Mo	98	1514.9	3.6	**0.5995**	0.027	4.5	ug/L	11	Standard
	Ag	107	86.3	3.5	**0.0012**	0.001	48.9	ug/L	68	Standard
	Cd	111	45.4	6.6	**0.0152**	0.002	11.0	mg/L	2	Standard
	Cd	114	119.3	31.2	**0.0227**	0.008	33.9	ug/L	12	Standard
>	**In**	115	551838.0	2.1				ug/L	587441	Standard
	Sn	118	896.7	9.3	**0.0688**	0.017	24.4	ug/L	497	Standard
	Sb	123	476.3	5.5	**0.1045**	0.008	7.8	ug/L	78	Standard
	Ba	135	51101.0	1.7	**26.5634**	0.162	0.6	ug/L	24	Standard
	Ce	140	4162.2	9.3				ug/L	25	Standard
>	**Tb**	159	892953.9	1.2				ug/L	939701	Standard
	Ho	165	111.7	20.2				ug/L	10	Standard
	Tl	203	71.0	9.8	**0.0072**	0.001	12.5	ug/L	4	Standard
	Tl	205	50.0	10.0	**0.0075**	0.001	10.0	ug/L	2	Standard
	Pb	206	834.7	3.9	**0.0918**	0.006	7.0	ug/L	240	Standard
	Pb	207	668.0	3.0	**0.0768**	0.004	5.0	ug/L	198	Standard
	Pb	208	2813.8	1.3	**0.0825**	0.002	2.7	ug/L	883	Standard
	U	238	2276.2	2.7	**0.3405**	0.008	2.4	ug/L	3	Standard
>	**Bi**	209	502274.9	0.4				ug/L	523984	Standard

Sample ID: L1606098617
Report Date/Time: Wednesday, June 29, 2016 13:53:42
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**21.6273**	37.451	173.2	mg/L	0	Standard
Mg	24	11037.4	4.7	**16.0863**	0.689	4.3	mg/L	37	Standard
K	39	470.0	10.1	**1.0642**	0.126	11.9	mg/L	38	Standard
Ca	43	383.3	8.5	**24.9281**	2.309	9.3	mg/L	10	Standard
Fe	54	307.3	17.1	**0.1571**	0.051	32.4	mg/L	179	Standard
Fe	57	490.0	5.4	**0.8143**	0.108	13.3	mg/L	335	Standard
Sc-1	45	53828.8	0.6				mg/L	61029	Standard
Cl	35	29673.0	1.2				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	4564.0	9.2				ug/L	1547	Standard
P	31	17653.6	5.5				ug/L	15286	Standard
S	34	1876.8	5.1				ug/L	1572	Standard
Sr	88	1206.7	6.6				ug/L	73	Standard
C	12	186.7	18.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	173.0	22.9				mg/L	6	Standard
Ho-1	165	111.7	20.2				mg/L	10	Standard
Er	166	76.7	30.1				mg/L	13	Standard
I	127	8810.9	16.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		127.313	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.958	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098617
Report Date/Time: Wednesday, June 29, 2016 13:53:42
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	93.939
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	95.857
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098619
Sample Date/Time: Wednesday, June 29, 2016 13:54:36
Number of Replicates: 3
Autosampler Position: 326
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96013.3	1.6				ug/L	74699	Standard
	Be	9	225.0	13.5	**0.2062**	0.024	11.6	ug/L	8	Standard
	Al	27	3833929.3	1.4	**20.2847**	0.231	1.1	ug/L	1493	Standard
	Sc	45	56757.9	4.8				ug/L	61029	Standard
	Ti	47	657.0	3.0	**1.6551**	0.023	1.4	ug/L	35	Standard
	V	51	6468.0	2.1	**0.6763**	0.022	3.3	ug/L	1435	Standard
	Cr	52	12143.2	0.5	**0.6446**	0.043	6.7	ug/L	7635	Standard
	Cr	53	1533.4	8.3	**1.3595**	0.147	10.8	ug/L	363	Standard
	Mn	55	41319.9	1.1	**5.8936**	0.094	1.6	ug/L	747	Standard
	Co	59	1041.4	1.4	**0.1233**	0.005	3.8	ug/L	234	Standard
	Ni	60	1381.7	2.6	**0.6272**	0.001	0.2	ug/L	215	Standard
	Cu	65	22006.0	2.1	**11.9309**	0.100	0.8	ug/L	104	Standard
	Zn	66	12936.5	1.8	**11.6003**	0.101	0.9	ug/L	363	Standard
>	**Ge**	72	486496.6	2.4				ug/L	495142	Standard
	As	75	1429.3	4.7	**1.3578**	0.030	2.2	ug/L	-50	Standard
	Se	82	368.2	2.0	**3.2780**	0.104	3.2	ug/L	14	Standard
	Se-1	77	166.3	2.4	**1.6971**	0.106	6.2	ug/L	42	Standard
>	**Ga**	71	256.7	8.8				mg/L	42	Standard
	Rb	85	6678.1	5.5				ug/L	17	Standard
	Y	89	411615.1	2.5				ug/L	417098	Standard
>	**Rh**	103	45.0	40.1				ug/L	13	Standard
	Mo	98	2416.6	3.5	**0.9418**	0.055	5.8	ug/L	11	Standard
	Ag	107	106.0	14.6	**0.0040**	0.002	49.0	ug/L	68	Standard
	Cd	111	306.3	1.8	**0.1451**	0.002	1.4	mg/L	2	Standard
	Cd	114	658.6	9.3	**0.1319**	0.011	8.6	ug/L	12	Standard
>	**In**	115	562685.3	3.0				ug/L	587441	Standard
	Sn	118	1296.7	5.8	**0.1370**	0.009	6.5	ug/L	497	Standard
	Sb	123	1866.5	3.1	**0.4132**	0.007	1.8	ug/L	78	Standard
	Ba	135	49124.5	1.8	**25.0474**	0.324	1.3	ug/L	24	Standard
	Ce	140	4487.3	4.7				ug/L	25	Standard
>	**Tb**	159	930596.3	3.6				ug/L	939701	Standard
	Ho	165	203.3	14.2				ug/L	10	Standard
	Tl	203	145.3	15.0	**0.0163**	0.002	14.2	ug/L	4	Standard
	Tl	205	131.7	4.4	**0.0191**	0.000	1.8	ug/L	2	Standard
	Pb	206	1800.8	1.2	**0.2755**	0.007	2.5	ug/L	240	Standard
	Pb	207	1451.7	2.6	**0.2445**	0.002	0.8	ug/L	198	Standard
	Pb	208	5832.1	0.4	**0.2459**	0.009	3.7	ug/L	883	Standard
	U	238	34205.2	1.0	**5.1077**	0.154	3.0	ug/L	3	Standard
>	**Bi**	209	509656.0	3.1				ug/L	523984	Standard

Sample ID: L1606098619
Report Date/Time: Wednesday, June 29, 2016 13:56:53
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	92209.8	1.7	**128.1856**	8.381	6.5	mg/L	37	Standard
K	39	298.3	3.5	**0.5957**	0.033	5.6	mg/L	38	Standard
Ca	43	80.0	6.3	**2.9252**	0.551	18.8	mg/L	10	Standard
Fe	54	254.0	14.7	**0.0930**	0.030	32.4	mg/L	179	Standard
Fe	57	333.3	5.3	**0.1453**	0.073	50.0	mg/L	335	Standard
> **Sc-1**	45	56757.9	4.8				mg/L	61029	Standard
Cl	35	31152.7	1.5				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	53109.7	4.4				ug/L	1547	Standard
P	31	17403.3	2.0				ug/L	15286	Standard
S	34	1883.5	7.3				ug/L	1572	Standard
Sr	88	213.3	3.6				ug/L	73	Standard
C	12	256.7	12.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	288.4	16.8				mg/L	6	Standard
Ho-1	165	203.3	14.2				mg/L	10	Standard
Er	166	173.3	23.3				mg/L	13	Standard
I	127	470221.1	5.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		128.534	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.254	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098619
Report Date/Time: Wednesday, June 29, 2016 13:56:53
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	95.786
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	97.265
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098625
Sample Date/Time: Wednesday, June 29, 2016 13:57:48
Number of Replicates: 3
Autosampler Position: 327
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	102613.0	0.8				ug/L	74699	Standard
	Be	9	78.3	9.8	**0.0724**	0.006	8.8	ug/L	8	Standard
	Al	27	33905192.3	1.5	**167.9500**	3.413	2.0	ug/L	1493	Standard
	Sc	45	63494.4	0.7				ug/L	61029	Standard
	Ti	47	6149.2	2.1	**15.7581**	0.447	2.8	ug/L	35	Standard
	V	51	12969.6	1.6	**1.4837**	0.027	1.8	ug/L	1435	Standard
	Cr	52	17892.2	0.6	**1.3372**	0.031	2.3	ug/L	7635	Standard
	Cr	53	3548.7	0.7	**3.4248**	0.009	0.3	ug/L	363	Standard
	Mn	55	3105877.1	1.7	**436.1197**	10.576	2.4	ug/L	747	Standard
	Co	59	6022.9	2.6	**0.8547**	0.030	3.5	ug/L	234	Standard
	Ni	60	6318.7	1.1	**3.1764**	0.053	1.7	ug/L	215	Standard
	Cu	65	5119.5	1.0	**2.5860**	0.011	0.4	ug/L	104	Standard
	Zn	66	19836.6	0.9	**17.3044**	0.276	1.6	ug/L	363	Standard
>	**Ge**	72	506357.7	0.8				ug/L	495142	Standard
	As	75	1090.3	6.9	**1.0117**	0.071	7.0	ug/L	-50	Standard
	Se	82	334.6	3.6	**2.8439**	0.125	4.4	ug/L	14	Standard
	Se-1	77	270.0	7.0	**2.9935**	0.272	9.1	ug/L	42	Standard
>	**Ga**	71	1363.4	8.0				mg/L	42	Standard
	Rb	85	42725.2	0.9				ug/L	17	Standard
	Y	89	453938.8	1.4				ug/L	417098	Standard
>	**Rh**	103	198.3	19.1				ug/L	13	Standard
	Mo	98	1348.4	1.8	**0.4789**	0.005	1.1	ug/L	11	Standard
	Ag	107	259.3	4.5	**0.0252**	0.002	7.6	ug/L	68	Standard
	Cd	111	351.0	2.3	**0.1529**	0.002	1.6	mg/L	2	Standard
	Cd	114	808.7	5.9	**0.1490**	0.010	6.6	ug/L	12	Standard
>	**In**	115	613191.6	0.9				ug/L	587441	Standard
	Sn	118	795.0	1.9	**0.0357**	0.003	9.4	ug/L	497	Standard
	Sb	123	319.9	2.9	**0.0614**	0.002	3.0	ug/L	78	Standard
	Ba	135	112678.5	1.7	**52.7237**	0.702	1.3	ug/L	24	Standard
	Ce	140	37260.8	2.0				ug/L	25	Standard
>	**Tb**	159	1001354.7	0.4				ug/L	939701	Standard
	Ho	165	1051.7	3.8				ug/L	10	Standard
	Tl	203	322.7	6.7	**0.0353**	0.003	7.6	ug/L	4	Standard
	Tl	205	283.3	17.7	**0.0376**	0.006	16.6	ug/L	2	Standard
	Pb	206	8100.8	1.1	**1.3649**	0.025	1.8	ug/L	240	Standard
	Pb	207	6526.4	1.1	**1.2323**	0.027	2.2	ug/L	198	Standard
	Pb	208	27115.2	0.7	**1.2953**	0.018	1.4	ug/L	883	Standard
	U	238	26591.0	2.4	**3.6542**	0.116	3.2	ug/L	3	Standard
>	**Bi**	209	553363.8	1.0				ug/L	523984	Standard

Sample ID: L1606098625
Report Date/Time: Wednesday, June 29, 2016 14:00:05
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**55.1859**	55.233	100.1	mg/L	0	Standard
Mg	24	8027.1	0.4	**9.8947**	0.087	0.9	mg/L	37	Standard
K	39	523.3	14.3	**0.9970**	0.152	15.2	mg/L	38	Standard
Ca	43	861.7	11.2	**49.8050**	6.018	12.1	mg/L	10	Standard
Fe	54	1108.4	4.4	**0.7648**	0.035	4.5	mg/L	179	Standard
Fe	57	975.0	10.7	**2.1112**	0.341	16.2	mg/L	335	Standard
> **Sc-1**	45	63494.4	0.7				mg/L	61029	Standard
Cl	35	33805.7	1.3				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	35418.0	3.4				ug/L	1547	Standard
P	31	20113.3	1.1				ug/L	15286	Standard
S	34	2053.5	4.6				ug/L	1572	Standard
Sr	88	896.7	5.2				ug/L	73	Standard
C	12	130.0	35.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	1595.5	8.0				mg/L	6	Standard
Ho-1	165	1051.7	3.8				mg/L	10	Standard
Er	166	1000.0	4.0				mg/L	13	Standard
I	127	43091.5	8.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		137.369	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		102.265	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098625
Report Date/Time: Wednesday, June 29, 2016 14:00:05
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	104.384
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	105.607
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098625
Report Date/Time: Wednesday, June 29, 2016 14:00:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098627
Sample Date/Time: Wednesday, June 29, 2016 14:00:59
Number of Replicates: 3
Autosampler Position: 328
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98891.7	1.0				ug/L	74699	Standard
	Be	9	638.3	5.3	**0.5552**	0.024	4.3	ug/L	8	Standard
	Al	27	63228292.3	2.4	**324.9422**	4.583	1.4	ug/L	1493	Standard
	Sc	45	59185.4	3.1				ug/L	61029	Standard
	Ti	47	12099.8	1.5	**31.9570**	0.495	1.6	ug/L	35	Standard
	V	51	31610.6	0.8	**4.0089**	0.089	2.2	ug/L	1435	Standard
	Cr	52	29501.7	2.0	**2.9745**	0.066	2.2	ug/L	7635	Standard
	Cr	53	4110.6	4.5	**4.1359**	0.142	3.4	ug/L	363	Standard
	Mn	55	939161.3	0.8	**135.3408**	0.783	0.6	ug/L	747	Standard
	Co	59	11927.7	1.5	**1.7769**	0.013	0.7	ug/L	234	Standard
	Ni	60	10174.7	1.2	**5.3344**	0.025	0.5	ug/L	215	Standard
	Cu	65	63138.6	0.6	**33.9717**	0.415	1.2	ug/L	104	Standard
	Zn	66	96624.1	0.5	**88.3979**	1.021	1.2	ug/L	363	Standard
>	**Ge**	72	492984.7	1.4				ug/L	495142	Standard
	As	75	1814.9	1.7	**1.6870**	0.014	0.9	ug/L	-50	Standard
	Se	82	95.5	0.5	**0.7363**	0.016	2.2	ug/L	14	Standard
	Se-1	77	118.3	17.1	**1.0027**	0.256	25.5	ug/L	42	Standard
>	**Ga**	71	3340.4	1.3				mg/L	42	Standard
	Rb	85	72434.4	1.9				ug/L	17	Standard
	Y	89	474043.5	0.9				ug/L	417098	Standard
>	**Rh**	103	386.7	16.4				ug/L	13	Standard
	Mo	98	2396.5	0.7	**0.9091**	0.018	1.9	ug/L	11	Standard
	Ag	107	338.3	4.7	**0.0399**	0.003	6.7	ug/L	68	Standard
	Cd	111	101.2	9.9	**0.0414**	0.005	12.6	mg/L	2	Standard
	Cd	114	230.2	21.2	**0.0435**	0.009	20.9	ug/L	12	Standard
>	**In**	115	577417.6	1.2				ug/L	587441	Standard
	Sn	118	1793.4	6.6	**0.2178**	0.023	10.5	ug/L	497	Standard
	Sb	123	600.3	4.8	**0.1267**	0.008	6.1	ug/L	78	Standard
	Ba	135	70114.5	1.5	**34.8357**	0.414	1.2	ug/L	24	Standard
	Ce	140	125576.7	1.8				ug/L	25	Standard
>	**Tb**	159	953741.7	1.0				ug/L	939701	Standard
	Ho	165	3755.5	3.8				ug/L	10	Standard
	Tl	203	452.7	12.3	**0.0539**	0.006	11.6	ug/L	4	Standard
	Tl	205	385.0	10.1	**0.0548**	0.005	9.0	ug/L	2	Standard
	Pb	206	12183.9	0.5	**2.2443**	0.020	0.9	ug/L	240	Standard
	Pb	207	9920.2	0.8	**2.0519**	0.018	0.9	ug/L	198	Standard
	Pb	208	42107.7	1.6	**2.2055**	0.026	1.2	ug/L	883	Standard
	U	238	36138.4	1.7	**5.3249**	0.032	0.6	ug/L	3	Standard
>	**Bi**	209	516181.2	1.2				ug/L	523984	Standard

Sample ID: L1606098627
Report Date/Time: Wednesday, June 29, 2016 14:03:16
Page 1

Approved: July 04, 2016

Na	23	8.3	91.7	**99.8498**	90.493	90.6	mg/L	0	Standard
Mg	24	9524.7	1.9	**12.6179**	0.315	2.5	mg/L	37	Standard
K	39	723.4	10.8	**1.5367**	0.198	12.9	mg/L	38	Standard
Ca	43	1308.4	8.3	**82.6186**	4.814	5.8	mg/L	10	Standard
Fe	54	3102.8	2.2	**2.5749**	0.034	1.3	mg/L	179	Standard
Fe	57	1665.1	3.0	**4.7616**	0.022	0.5	mg/L	335	Standard
> **Sc-1**	45	59185.4	3.1				mg/L	61029	Standard
Cl	35	38114.5	0.9				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	13185.7	5.6				ug/L	1547	Standard
P	31	20465.5	3.4				ug/L	15286	Standard
S	34	1960.1	7.7				ug/L	1572	Standard
Sr	88	1940.1	7.4				ug/L	73	Standard
C	12	216.7	23.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	3	Standard
Dy	164	5933.6	2.4				mg/L	6	Standard
Ho-1	165	3755.5	3.8				mg/L	10	Standard
Er	166	3430.4	0.6				mg/L	13	Standard
I	127	31850.1	1.9				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		132.387	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		99.564	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606098627
Report Date/Time: Wednesday, June 29, 2016 14:03:16
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	98.294
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	98.511
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098627
Report Date/Time: Wednesday, June 29, 2016 14:03:16
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 14:04:12
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	95022.8	6.4				ug/L	74699	Standard
	Be	9	57413.6	2.5	**51.3575**	1.980	3.9	ug/L	8	Standard
	Al	27	7949888.2	1.3	**42.6051**	2.208	5.2	ug/L	1493	Standard
	Sc	45	58088.0	4.8				ug/L	61029	Standard
	Ti	47	37651.4	1.9	**100.2698**	3.249	3.2	ug/L	35	Standard
	V	51	404346.6	1.8	**53.8882**	0.844	1.6	ug/L	1435	Standard
	Cr	52	381698.1	1.1	**51.0223**	1.642	3.2	ug/L	7635	Standard
	Cr	53	47042.9	0.3	**51.1128**	1.842	3.6	ug/L	363	Standard
	Mn	55	357326.2	3.2	**51.6830**	0.417	0.8	ug/L	747	Standard
	Co	59	371699.6	2.5	**56.8155**	0.522	0.9	ug/L	234	Standard
	Ni	60	94541.7	1.4	**50.9011**	1.597	3.1	ug/L	215	Standard
	Cu	65	94983.8	1.5	**51.4773**	2.133	4.1	ug/L	104	Standard
	Zn	66	56900.0	0.7	**52.1829**	1.541	3.0	ug/L	363	Standard
>	Ge	72	490310.3	3.4				ug/L	495142	Standard
	As	75	57867.9	1.8	**52.2509**	0.921	1.8	ug/L	-50	Standard
	Se	82	5817.8	2.4	**53.4067**	0.847	1.6	ug/L	14	Standard
	Se-1	77	4012.5	1.4	**54.8814**	1.095	2.0	ug/L	42	Standard
>	Ga	71	80.0	6.3				mg/L	42	Standard
	Rb	85	831.7	8.5				ug/L	17	Standard
	Y	89	423000.5	2.4				ug/L	417098	Standard
>	Rh	103	26.7	57.3				ug/L	13	Standard
	Mo	98	279510.1	1.2	**103.0732**	1.481	1.4	ug/L	11	Standard
	Ag	107	337025.9	0.7	**50.8187**	0.653	1.3	ug/L	68	Standard
	Cd	111	104719.9	1.6	**49.2197**	0.492	1.0	mg/L	2	Standard
	Cd	114	254484.5	1.8	**48.7789**	1.357	2.8	ug/L	12	Standard
>	In	115	597695.8	1.6				ug/L	587441	Standard
	Sn	118	297589.5	3.3	**49.9660**	0.932	1.9	ug/L	497	Standard
	Sb	123	237800.4	0.6	**50.0611**	0.480	1.0	ug/L	78	Standard
	Ba	135	98785.4	1.3	**47.4247**	0.676	1.4	ug/L	24	Standard
	Ce	140	118.3	6.5				ug/L	25	Standard
>	Tb	159	962162.8	1.8				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	431721.7	1.4	**50.5906**	0.299	0.6	ug/L	4	Standard
	Tl	205	386667.7	1.8	**52.2055**	0.331	0.6	ug/L	2	Standard
	Pb	206	286495.7	0.6	**51.7458**	1.032	2.0	ug/L	240	Standard
	Pb	207	254256.8	1.2	**51.6830**	0.707	1.4	ug/L	198	Standard
	Pb	208	993731.8	1.4	**51.1193**	0.294	0.6	ug/L	883	Standard
	U	238	371822.3	1.0	**52.1805**	0.682	1.3	ug/L	3	Standard
>	Bi	209	542470.6	1.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 14:06:29
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**20.6892**	35.826	173.2	mg/L	0	Standard
Mg	24	4115.6	5.8	**5.5148**	0.066	1.2	mg/L	37	Standard
K	39	1971.8	1.5	**4.4711**	0.214	4.8	mg/L	38	Standard
Ca	43	71.7	16.1	**2.2502**	0.891	39.6	mg/L	10	Standard
Fe	54	5897.1	1.5	**5.1229**	0.272	5.3	mg/L	179	Standard
Fe	57	1798.4	2.6	**5.3592**	0.387	7.2	mg/L	335	Standard
> **Sc-1**	45	58088.0	4.8				mg/L	61029	Standard
Cl	35	26443.8	4.7				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	1943.5	9.9				ug/L	1547	Standard
P	31	18503.0	1.3				ug/L	15286	Standard
S	34	2213.5	3.2				ug/L	1572	Standard
Sr	88	76.7	10.0				ug/L	73	Standard
C	12	120.0	43.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	2.9	217.0				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	3483.8	25.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	102.715		
Al	27	85.210		
Sc	45			
Ti	47	100.270		
V	51	107.776		
Cr	52	102.045		
Cr	53			
Mn	55	103.366		
Co	59	113.631		
Ni	60	101.802		
Cu	65	102.955		
Zn	66	104.366		
> Ge	72		99.024	
As	75	104.502		
Se	82	106.813		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 14:06:29
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	103.073		
Ag	107	101.637		
Cd	111	98.439		
Cd	114			
In	115		101.746	
Sn	118	99.932		
Sb	123	100.122		
Ba	135	94.849		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.181		
Tl	205			
Pb	206			
Pb	207			
Pb	208	102.239		
U	238	104.361		
Bi	209		103.528	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 14:06:29
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 14:07:23
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87483.9	4.7				ug/L	74699	Standard
	Be	9	21.7	96.1	**0.0292**	0.021	73.0	ug/L	8	Standard
	Al	27	18230.8	94.8	**0.0950**	0.107	112.2	ug/L	1493	Standard
	Sc	45	56471.8	3.5				ug/L	61029	Standard
	Ti	47	36.0	46.7	**-0.0098**	0.050	510.8	ug/L	35	Standard
	V	51	1334.8	2.3	**-0.0125**	0.010	80.4	ug/L	1435	Standard
	Cr	52	6684.1	1.9	**-0.0897**	0.031	35.0	ug/L	7635	Standard
	Cr	53	393.3	6.4	**0.1115**	0.043	38.6	ug/L	363	Standard
	Mn	55	1346.4	57.0	**0.0546**	0.123	224.6	ug/L	747	Standard
	Co	59	278.3	17.3	**0.0065**	0.009	139.3	ug/L	234	Standard
	Ni	60	177.0	5.9	**-0.0261**	0.004	17.0	ug/L	215	Standard
	Cu	65	145.0	10.8	**-0.0230**	0.009	41.3	ug/L	104	Standard
	Zn	66	399.3	2.4	**-0.0757**	0.021	28.0	ug/L	363	Standard
>	**Ge**	72	478426.6	3.3				ug/L	495142	Standard
	As	75	-39.3	49.9	**0.0231**	0.019	81.6	ug/L	-50	Standard
	Se	82	16.5	11.7	**0.0181**	0.023	124.9	ug/L	14	Standard
	Se-1	77	38.7	26.9	**-0.0773**	0.131	170.1	ug/L	42	Standard
>	**Ga**	71	66.7	15.6				mg/L	42	Standard
	Rb	85	18.3	41.7				ug/L	17	Standard
	Y	89	416242.9	2.8				ug/L	417098	Standard
>	**Rh**	103	15.0	57.7				ug/L	13	Standard
	Mo	98	272.1	23.3	**0.0976**	0.024	24.4	ug/L	11	Standard
	Ag	107	106.0	6.8	**0.0036**	0.001	28.2	ug/L	68	Standard
	Cd	111	10.2	22.2	**-0.0030**	0.001	37.6	mg/L	2	Standard
	Cd	114	31.6	53.8	**0.0041**	0.003	80.7	ug/L	12	Standard
>	**In**	115	580179.3	0.7				ug/L	587441	Standard
	Sn	118	746.7	16.1	**0.0348**	0.022	62.5	ug/L	497	Standard
	Sb	123	368.4	24.4	**0.0758**	0.020	26.4	ug/L	78	Standard
	Ba	135	41.0	10.6	**0.0043**	0.002	51.4	ug/L	24	Standard
	Ce	140	38.3	61.6				ug/L	25	Standard
>	**Tb**	159	945404.9	0.5				ug/L	939701	Standard
	Ho	165	6.7	43.3				ug/L	10	Standard
	Tl	203	65.7	8.7	**0.0061**	0.001	8.8	ug/L	4	Standard
	Tl	205	95.0	45.9	**0.0133**	0.006	45.0	ug/L	2	Standard
	Pb	206	276.0	7.9	**-0.0202**	0.003	15.8	ug/L	240	Standard
	Pb	207	258.0	12.6	**-0.0163**	0.006	36.3	ug/L	198	Standard
	Pb	208	1134.3	12.6	**-0.0143**	0.008	56.7	ug/L	883	Standard
	U	238	91.0	75.8	**0.0087**	0.010	115.9	ug/L	3	Standard
>	**Bi**	209	530301.2	1.8				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 14:09:40
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	81.1	**0.0007**	0.054	7923.6	mg/L	37	Standard
K	39	40.0	45.1	**-0.0194**	0.039	202.6	mg/L	38	Standard
Ca	43	28.3	20.4	**-0.5841**	0.427	73.1	mg/L	10	Standard
Fe	54	129.2	39.5	**-0.0191**	0.049	258.2	mg/L	179	Standard
Fe	57	241.7	27.6	**-0.1803**	0.270	149.7	mg/L	335	Standard
> **Sc-1**	45	56471.8	3.5				mg/L	61029	Standard
Cl	35	26027.0	1.3				ug/L	20195	Standard
Kr	83	3.0	57.7				ug/L	4	Standard
Br	81	1783.4	6.5				ug/L	1547	Standard
P	31	18152.5	2.7				ug/L	15286	Standard
S	34	2168.5	1.2				ug/L	1572	Standard
Sr	88	56.7	27.0				ug/L	73	Standard
C	12	93.3	24.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	6.5	86.7				mg/L	6	Standard
Ho-1	165	6.7	43.3				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	2606.9	21.7				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		96.624	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 14:09:40
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	98.764	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	101.206	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 14:09:40
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098629
Sample Date/Time: Wednesday, June 29, 2016 14:10:36
Number of Replicates: 3
Autosampler Position: 329
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98303.0	2.6				ug/L	74699	Standard
	Be	9	480.0	19.8	**0.4216**	0.078	18.5	ug/L	8	Standard
	Al	27	74581592.0	2.0	**385.7852**	11.899	3.1	ug/L	1493	Standard
	Sc	45	57877.0	1.4				ug/L	61029	Standard
	Ti	47	6572.4	2.9	**18.4374**	0.321	1.7	ug/L	35	Standard
	V	51	8784.7	0.2	**1.0487**	0.019	1.8	ug/L	1435	Standard
	Cr	52	16164.0	1.7	**1.3092**	0.018	1.4	ug/L	7635	Standard
	Cr	53	2105.1	0.9	**2.1070**	0.052	2.5	ug/L	363	Standard
	Mn	55	1145494.2	1.1	**175.8506**	2.171	1.2	ug/L	747	Standard
	Co	59	6198.3	3.2	**0.9668**	0.035	3.6	ug/L	234	Standard
	Ni	60	8328.3	2.3	**4.6339**	0.069	1.5	ug/L	215	Standard
	Cu	65	39558.9	1.3	**22.6316**	0.139	0.6	ug/L	104	Standard
	Zn	66	37507.0	2.0	**36.2726**	0.401	1.1	ug/L	363	Standard
>	**Ge**	72	462898.1	1.3				ug/L	495142	Standard
	As	75	729.1	4.9	**0.7559**	0.035	4.6	ug/L	-50	Standard
	Se	82	97.9	9.3	**0.8168**	0.100	12.2	ug/L	14	Standard
	Se-1	77	92.7	14.7	**0.7365**	0.213	29.0	ug/L	42	Standard
>	**Ga**	71	1238.4	2.6				mg/L	42	Standard
	Rb	85	36981.8	1.6				ug/L	17	Standard
	Y	89	423750.9	0.9				ug/L	417098	Standard
>	**Rh**	103	1170.0	7.1				ug/L	13	Standard
	Mo	98	1333.1	1.3	**0.5111**	0.005	0.9	ug/L	11	Standard
	Ag	107	159.7	2.8	**0.0124**	0.001	8.3	ug/L	68	Standard
	Cd	111	50.7	12.0	**0.0171**	0.003	15.0	mg/L	2	Standard
	Cd	114	154.3	14.1	**0.0289**	0.004	13.0	ug/L	12	Standard
>	**In**	115	568573.6	2.1				ug/L	587441	Standard
	Sn	118	2385.2	8.8	**0.3269**	0.032	9.8	ug/L	497	Standard
	Sb	123	701.5	2.0	**0.1511**	0.004	2.9	ug/L	78	Standard
	Ba	135	49620.1	0.4	**25.0380**	0.447	1.8	ug/L	24	Standard
	Ce	140	69524.2	1.7				ug/L	25	Standard
>	**Tb**	159	941672.2	2.5				ug/L	939701	Standard
	Ho	165	1511.7	2.5				ug/L	10	Standard
	Tl	203	215.3	8.2	**0.0254**	0.002	6.7	ug/L	4	Standard
	Tl	205	171.7	4.4	**0.0251**	0.001	2.6	ug/L	2	Standard
	Pb	206	5244.6	2.1	**0.9485**	0.002	0.2	ug/L	240	Standard
	Pb	207	4168.9	1.9	**0.8423**	0.008	0.9	ug/L	198	Standard
	Pb	208	17565.0	0.7	**0.8990**	0.013	1.4	ug/L	883	Standard
	U	238	17410.3	1.3	**2.6228**	0.019	0.7	ug/L	3	Standard
>	**Bi**	209	504494.7	1.9				ug/L	523984	Standard

Sample ID: L1606098629
Report Date/Time: Wednesday, June 29, 2016 14:12:53
Page 1

Approved: July 04, 2016

Na	23	13.3	78.1	**161.6266**	126.012	78.0	mg/L	0	Standard
Mg	24	27959.1	0.6	**37.9939**	0.282	0.7	mg/L	37	Standard
K	39	733.4	5.5	**1.5946**	0.079	5.0	mg/L	38	Standard
Ca	43	1693.4	2.8	**110.2950**	4.030	3.7	mg/L	10	Standard
Fe	54	1507.3	3.7	**1.2090**	0.035	2.9	mg/L	179	Standard
Fe	57	1286.7	5.0	**3.5362**	0.177	5.0	mg/L	335	Standard
> **Sc-1**	45	57877.0	1.4				mg/L	61029	Standard
Cl	35	39812.2	0.7				ug/L	20195	Standard
Kr	83	2.3	24.7				ug/L	4	Standard
Br	81	15855.0	0.8				ug/L	1547	Standard
P	31	19874.7	1.1				ug/L	15286	Standard
S	34	1861.8	8.3				ug/L	1572	Standard
Sr	88	5489.3	0.8				ug/L	73	Standard
C	12	143.3	29.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	2426.6	8.1				mg/L	6	Standard
Ho-1	165	1511.7	2.5				mg/L	10	Standard
Er	166	1466.7	15.1				mg/L	13	Standard
I	127	114010.2	6.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		131.599	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.488	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098629
Report Date/Time: Wednesday, June 29, 2016 14:12:53
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	96.788
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	96.280
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606098629
Report Date/Time: Wednesday, June 29, 2016 14:12:53
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098639
Sample Date/Time: Wednesday, June 29, 2016 14:13:47
Number of Replicates: 3
Autosampler Position: 330
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95709.8	2.9				ug/L	74699	Standard
	Be	9	26.7	92.5	**0.0314**	0.022	70.9	ug/L	8	Standard
	Al	27	47777.1	19.3	**0.2396**	0.041	17.2	ug/L	1493	Standard
	Sc	45	57199.5	2.6				ug/L	61029	Standard
	Ti	47	140.3	60.7	**0.2633**	0.227	86.2	ug/L	35	Standard
	V	51	2113.6	5.5	**0.0856**	0.015	17.8	ug/L	1435	Standard
	Cr	52	11519.7	1.7	**0.5382**	0.024	4.4	ug/L	7635	Standard
	Cr	53	731.7	8.2	**0.4658**	0.068	14.5	ug/L	363	Standard
	Mn	55	10046.3	0.4	**1.3019**	0.003	0.2	ug/L	747	Standard
	Co	59	277.3	4.9	**0.0049**	0.002	38.1	ug/L	234	Standard
	Ni	60	403.7	2.5	**0.0928**	0.006	6.6	ug/L	215	Standard
	Cu	65	2594.2	1.0	**1.2969**	0.011	0.8	ug/L	104	Standard
	Zn	66	5488.0	1.5	**4.5940**	0.065	1.4	ug/L	363	Standard
>	**Ge**	72	492746.6	0.5				ug/L	495142	Standard
	As	75	-23.9	159.7	**0.0381**	0.034	90.0	ug/L	-50	Standard
	Se	82	16.9	42.3	**0.0162**	0.065	398.3	ug/L	14	Standard
	Se-1	77	44.7	15.2	**-0.0080**	0.091	1136.4	ug/L	42	Standard
>	**Ga**	71	35.0	24.7				mg/L	42	Standard
	Rb	85	228.3	6.7				ug/L	17	Standard
	Y	89	426313.2	2.4				ug/L	417098	Standard
>	**Rh**	103	20.0	66.1				ug/L	13	Standard
	Mo	98	146.2	8.6	**0.0487**	0.005	10.8	ug/L	11	Standard
	Ag	107	84.7	10.6	**-0.0000**	0.001	3427.7	ug/L	68	Standard
	Cd	111	14.4	26.2	**-0.0011**	0.002	163.5	mg/L	2	Standard
	Cd	114	36.0	7.5	**0.0048**	0.001	11.2	ug/L	12	Standard
>	**In**	115	592771.3	1.1				ug/L	587441	Standard
	Sn	118	1206.7	2.0	**0.1101**	0.007	6.0	ug/L	497	Standard
	Sb	123	2747.6	2.7	**0.5790**	0.016	2.8	ug/L	78	Standard
	Ba	135	327.7	1.8	**0.1426**	0.002	1.3	ug/L	24	Standard
	Ce	140	215.0	46.7				ug/L	25	Standard
>	**Tb**	159	942434.8	1.8				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	30.7	28.5	**0.0019**	0.001	54.2	ug/L	4	Standard
	Tl	205	31.7	59.8	**0.0045**	0.003	57.0	ug/L	2	Standard
	Pb	206	347.7	8.5	**-0.0077**	0.005	63.4	ug/L	240	Standard
	Pb	207	296.0	3.8	**-0.0091**	0.002	22.4	ug/L	198	Standard
	Pb	208	1207.0	5.0	**-0.0113**	0.003	23.3	ug/L	883	Standard
	U	238	16.7	21.1	**-0.0021**	0.000	23.4	ug/L	3	Standard
>	**Bi**	209	536687.4	0.8				ug/L	523984	Standard

Sample ID: L1606098639
Report Date/Time: Wednesday, June 29, 2016 14:16:04
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	78.3	9.8	**0.0427**	0.008	19.8	mg/L	37	Standard
K	39	33.3	56.8	**-0.0347**	0.047	136.1	mg/L	38	Standard
Ca	43	38.3	27.2	**0.0551**	0.672	1218.0	mg/L	10	Standard
Fe	54	152.5	13.6	**-0.0001**	0.021	24429.9	mg/L	179	Standard
Fe	57	270.0	10.3	**-0.0961**	0.083	86.0	mg/L	335	Standard
> **Sc-1**	45	57199.5	2.6				mg/L	61029	Standard
Cl	35	30633.6	0.6				ug/L	20195	Standard
Kr	83	2.7	78.1				ug/L	4	Standard
Br	81	2093.5	12.2				ug/L	1547	Standard
P	31	17446.7	2.4				ug/L	15286	Standard
S	34	1971.8	2.9				ug/L	1572	Standard
Sr	88	48.3	26.0				ug/L	73	Standard
C	12	116.7	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	9.0	108.0				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	4304.0	50.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		128.128	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.516	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606098639
Report Date/Time: Wednesday, June 29, 2016 14:16:04
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.907
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.424
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606098641
Sample Date/Time: Wednesday, June 29, 2016 14:16:59
Number of Replicates: 3
Autosampler Position: 331
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87267.4	3.2				ug/L	74699	Standard
	Be	9	11.7	49.5	**0.0189**	0.005	28.8	ug/L	8	Standard
	Al	27	605511.0	1.7	**3.5149**	0.083	2.4	ug/L	1493	Standard
	Sc	45	54390.8	2.9				ug/L	61029	Standard
	Ti	47	93.0	12.4	**0.1442**	0.032	22.4	ug/L	35	Standard
	V	51	1794.1	5.9	**0.0496**	0.015	30.2	ug/L	1435	Standard
	Cr	52	10365.2	1.5	**0.4195**	0.024	5.7	ug/L	7635	Standard
	Cr	53	613.3	8.2	**0.3551**	0.058	16.5	ug/L	363	Standard
	Mn	55	4513.0	1.6	**0.5208**	0.012	2.3	ug/L	747	Standard
	Co	59	310.3	5.3	**0.0112**	0.002	21.0	ug/L	234	Standard
	Ni	60	490.0	6.7	**0.1461**	0.017	11.6	ug/L	215	Standard
	Cu	65	1018.0	2.6	**0.4608**	0.017	3.7	ug/L	104	Standard
	Zn	66	3968.5	2.5	**3.2944**	0.091	2.8	ug/L	363	Standard
>	**Ge**	72	479837.7	0.4				ug/L	495142	Standard
	As	75	-18.8	39.3	**0.0423**	0.007	15.9	ug/L	-50	Standard
	Se	82	16.1	20.6	**0.0139**	0.032	229.1	ug/L	14	Standard
	Se-1	77	51.0	14.8	**0.0980**	0.104	105.7	ug/L	42	Standard
>	**Ga**	71	46.7	22.3				mg/L	42	Standard
	Rb	85	2146.8	6.0				ug/L	17	Standard
	Y	89	405082.7	1.5				ug/L	417098	Standard
>	**Rh**	103	10.0	86.6				ug/L	13	Standard
	Mo	98	183.9	8.1	**0.0672**	0.006	9.5	ug/L	11	Standard
	Ag	107	85.0	8.5	**0.0009**	0.001	123.2	ug/L	68	Standard
	Cd	111	58.3	9.8	**0.0215**	0.003	12.3	mg/L	2	Standard
	Cd	114	162.0	10.0	**0.0312**	0.003	9.6	ug/L	12	Standard
>	**In**	115	556306.9	1.6				ug/L	587441	Standard
	Sn	118	983.4	6.5	**0.0831**	0.013	16.0	ug/L	497	Standard
	Sb	123	290.8	14.6	**0.0617**	0.011	17.2	ug/L	78	Standard
	Ba	135	1301.4	2.0	**0.6554**	0.007	1.1	ug/L	24	Standard
	Ce	140	1251.7	2.7				ug/L	25	Standard
>	**Tb**	159	894371.7	2.0				ug/L	939701	Standard
	Ho	165	23.3	61.9				ug/L	10	Standard
	Tl	203	26.7	21.3	**0.0015**	0.001	46.3	ug/L	4	Standard
	Tl	205	18.3	68.6	**0.0028**	0.002	63.1	ug/L	2	Standard
	Pb	206	368.7	5.0	**-0.0008**	0.004	496.5	ug/L	240	Standard
	Pb	207	301.3	5.5	**-0.0052**	0.004	82.3	ug/L	198	Standard
	Pb	208	1198.4	1.5	**-0.0089**	0.002	20.6	ug/L	883	Standard
	U	238	7.7	32.8	**-0.0033**	0.000	11.3	ug/L	3	Standard
>	**Bi**	209	513924.4	1.5				ug/L	523984	Standard

Sample ID: L1606098641
Report Date/Time: Wednesday, June 29, 2016 14:19:15
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	638.3	13.6	**0.8624**	0.151	17.5	mg/L	37	Standard
K	39	86.7	3.3	**0.1010**	0.012	12.3	mg/L	38	Standard
Ca	43	31.7	48.2	**-0.2871**	1.074	374.0	mg/L	10	Standard
Fe	54	129.1	21.5	**-0.0150**	0.029	192.4	mg/L	179	Standard
Fe	57	250.0	12.2	**-0.1181**	0.141	119.7	mg/L	335	Standard
Sc-1	45	54390.8	2.9				mg/L	61029	Standard
Cl	35	28123.4	3.1				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1706.8	4.2				ug/L	1547	Standard
P	31	17456.7	1.2				ug/L	15286	Standard
S	34	1960.1	4.3				ug/L	1572	Standard
Sr	88	86.7	3.3				ug/L	73	Standard
C	12	86.7	35.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	22.9	67.6				mg/L	6	Standard
Ho-1	165	23.3	61.9				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	6631.5	5.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		116.826	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		96.909	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606098641
Report Date/Time: Wednesday, June 29, 2016 14:19:15
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.700
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.080
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102602
Sample Date/Time: Wednesday, June 29, 2016 14:51:20
Number of Replicates: 3
Autosampler Position: 332
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98959.1	4.1				ug/L	74699	Standard
	Be	9	893.4	7.1	**0.7730**	0.024	3.1	ug/L	8	Standard
	Al	27	86838575.1	3.0	**446.1622**	5.675	1.3	ug/L	1493	Standard
	Sc	45	59841.3	1.2				ug/L	61029	Standard
	Ti	47	724.7	2.7	**1.9675**	0.086	4.4	ug/L	35	Standard
	V	51	-99401.9	27.0	**-14.4927**	3.863	26.7	ug/L	1435	Standard
	Cr	52	14286817.6	0.7	**2092.1702**	38.595	1.8	ug/L	7635	Standard
	Cr	53	1819264.9	1.0	**2136.8631**	30.527	1.4	ug/L	363	Standard
	Mn	55	9369393.8	2.3	**1460.7209**	18.166	1.2	ug/L	747	Standard
	Co	59	953653.0	0.8	**156.7618**	2.325	1.5	ug/L	234	Standard
	Ni	60	10094931.6	0.5	**5854.0954**	89.919	1.5	ug/L	215	Standard
	Cu	65	555284.9	0.4	**323.8508**	6.370	2.0	ug/L	104	Standard
	Zn	66	64094.6	1.0	**63.2545**	1.058	1.7	ug/L	363	Standard
>	**Ge**	72	456115.5	1.9				ug/L	495142	Standard
	As	75	3627.2	1.9	**3.5767**	0.135	3.8	ug/L	-50	Standard
	Se	82	1754.2	3.0	**17.2111**	0.289	1.7	ug/L	14	Standard
	Se-1	77	3450.1	4.6	**50.7001**	3.400	6.7	ug/L	42	Standard
>	**Ga**	71	493.3	9.8				mg/L	42	Standard
	Rb	85	15579.7	1.3				ug/L	17	Standard
	Y	89	510259.3	1.5				ug/L	417098	Standard
>	**Rh**	103	536.7	10.8				ug/L	13	Standard
	Mo	98	52406.3	2.2	**22.0721**	0.200	0.9	ug/L	11	Standard
	Ag	107	797.7	2.2	**0.1246**	0.004	3.1	ug/L	68	Standard
	Cd	111	4820.0	1.8	**2.5808**	0.025	1.0	mg/L	2	Standard
	Cd	114	11654.3	2.8	**2.5494**	0.039	1.5	ug/L	12	Standard
>	**In**	115	523117.8	1.3				ug/L	587441	Standard
	Sn	118	2368.5	5.4	**0.3606**	0.023	6.4	ug/L	497	Standard
	Sb	123	2240.5	1.3	**0.5347**	0.001	0.2	ug/L	78	Standard
	Ba	135	135783.1	0.9	**74.4848**	0.361	0.5	ug/L	24	Standard
	Ce	140	104687.1	1.1				ug/L	25	Standard
>	**Tb**	159	898543.7	0.7				ug/L	939701	Standard
	Ho	165	6684.8	2.2				ug/L	10	Standard
	Tl	203	178.0	4.1	**0.0225**	0.001	5.6	ug/L	4	Standard
	Tl	205	203.3	37.2	**0.0321**	0.012	38.4	ug/L	2	Standard
	Pb	206	1304.4	5.0	**0.2026**	0.016	8.1	ug/L	240	Standard
	Pb	207	1055.7	2.5	**0.1793**	0.004	2.5	ug/L	198	Standard
	Pb	208	4341.2	2.0	**0.1854**	0.004	2.0	ug/L	883	Standard
	U	238	4585.0	0.8	**0.7423**	0.007	0.9	ug/L	3	Standard
>	**Bi**	209	467454.5	1.3				ug/L	523984	Standard

Sample ID: L1606102602
Report Date/Time: Wednesday, June 29, 2016 14:53:37
Page 1

Approved: July 04, 2016

Na	23	5.0	0.0	**58.7812**	0.729	1.2	mg/L	0	Standard
Mg	24	188890.8	2.9	**248.6881**	10.353	4.2	mg/L	37	Standard
K	39	2185.2	4.4	**4.8106**	0.175	3.6	mg/L	38	Standard
Ca	43	1135.0	6.5	**70.5642**	3.873	5.5	mg/L	10	Standard
Fe	54	9878.2	4.8	**8.4031**	0.399	4.8	mg/L	179	Standard
Fe	57	3267.0	5.5	**10.2558**	0.748	7.3	mg/L	335	Standard
Sc-1	45	59841.3	1.2				mg/L	61029	Standard
Cl	35	28240.3	1.1				ug/L	20195	Standard
Kr	83	2.7	142.0				ug/L	4	Standard
Br	81	339155.5	2.6				ug/L	1547	Standard
P	31	29607.2	0.7				ug/L	15286	Standard
S	34	2378.5	1.9				ug/L	1572	Standard
Sr	88	1895.1	7.4				ug/L	73	Standard
C	12	120.0	28.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	86.7	35.3				mg/L	3	Standard
Dy	164	9299.8	2.7				mg/L	6	Standard
Ho-1	165	6684.8	2.2				mg/L	10	Standard
Er	166	6331.3	1.7				mg/L	13	Standard
I	127	153619.9	12.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.478	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.118	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606102602
Report Date/Time: Wednesday, June 29, 2016 14:53:37
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		89.050
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		89.212
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
V 51 Lower	V	51	

Sample ID: L1606102602
Report Date/Time: Wednesday, June 29, 2016 14:53:37
Page 3

Approved: July 04, 2016

Cr 52 Upper, S, EEE	Cr	52
Cr 53 Upper, S, EEE	Cr	53
Mn 55 Upper, S, EEE	Mn	55
Co 59 Upper, S, EEE	Co	59
Ni 60 Upper, S, EEE	Ni	60
Cu 65 Upper, S, EEE	Cu	65

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102604
Sample Date/Time: Wednesday, June 29, 2016 14:54:31
Number of Replicates: 3
Autosampler Position: 333
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82386.8	1.6				ug/L	74699	Standard
	Be	9	41.7	50.0	**0.0505**	0.021	41.9	ug/L	8	Standard
	Al	27	6418180.7	1.0	**39.5946**	1.017	2.6	ug/L	1493	Standard
	Sc	45	68191.6	3.4				ug/L	61029	Standard
	Ti	47	2431.5	5.8	**6.2763**	0.263	4.2	ug/L	35	Standard
	V	51	8932.1	2.8	**0.9822**	0.034	3.4	ug/L	1435	Standard
	Cr	52	19971.2	5.8	**1.6585**	0.113	6.8	ug/L	7635	Standard
	Cr	53	16018.5	1.2	**16.9336**	0.479	2.8	ug/L	363	Standard
	Mn	55	134583.1	0.5	**19.1035**	0.326	1.7	ug/L	747	Standard
	Co	59	5374.0	2.6	**0.7730**	0.008	1.1	ug/L	234	Standard
	Ni	60	7430.2	8.6	**3.8249**	0.274	7.2	ug/L	215	Standard
	Cu	65	1488.1	5.9	**0.6921**	0.035	5.0	ug/L	104	Standard
	Zn	66	5451.0	0.8	**4.5156**	0.102	2.3	ug/L	363	Standard
>	**Ge**	72	497254.0	1.7				ug/L	495142	Standard
	As	75	427.8	16.4	**0.4399**	0.063	14.3	ug/L	-50	Standard
	Se	82	161.6	5.9	**1.3278**	0.065	4.9	ug/L	14	Standard
	Se-1	77	695.3	4.1	**8.8552**	0.225	2.5	ug/L	42	Standard
>	**Ga**	71	420.0	2.4				mg/L	42	Standard
	Rb	85	8292.3	2.6				ug/L	17	Standard
	Y	89	430205.9	2.4				ug/L	417098	Standard
>	**Rh**	103	40.0	12.5				ug/L	13	Standard
	Mo	98	177.9	8.2	**0.0624**	0.005	8.3	ug/L	11	Standard
	Ag	107	97.3	5.7	**0.0023**	0.001	38.0	ug/L	68	Standard
	Cd	111	330.0	5.3	**0.1533**	0.008	5.4	mg/L	2	Standard
	Cd	114	777.4	4.0	**0.1527**	0.007	4.3	ug/L	12	Standard
>	**In**	115	575297.3	2.0				ug/L	587441	Standard
	Sn	118	1836.8	3.5	**0.2263**	0.005	2.3	ug/L	497	Standard
	Sb	123	541.3	2.5	**0.1143**	0.005	4.2	ug/L	78	Standard
	Ba	135	259403.2	1.0	**129.4148**	1.250	1.0	ug/L	24	Standard
	Ce	140	48153.1	1.6				ug/L	25	Standard
>	**Tb**	159	953820.3	0.6				ug/L	939701	Standard
	Ho	165	763.4	3.9				ug/L	10	Standard
	Tl	203	147.7	11.5	**0.0162**	0.002	12.5	ug/L	4	Standard
	Tl	205	135.0	25.9	**0.0192**	0.005	26.2	ug/L	2	Standard
	Pb	206	2333.8	1.6	**0.3679**	0.004	1.0	ug/L	240	Standard
	Pb	207	1897.5	1.5	**0.3318**	0.006	1.8	ug/L	198	Standard
	Pb	208	7911.8	1.2	**0.3500**	0.006	1.7	ug/L	883	Standard
	U	238	272.0	3.5	**0.0353**	0.002	4.3	ug/L	3	Standard
>	**Bi**	209	521315.2	0.9				ug/L	523984	Standard

Sample ID: L1606102604
Report Date/Time: Wednesday, June 29, 2016 14:56:48
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	45454.8	3.6	**52.4459**	0.539	1.0	mg/L	37	Standard
K	39	88.3	3.3	**0.0609**	0.011	18.4	mg/L	38	Standard
Ca	43	100.0	8.7	**3.1440**	0.659	21.0	mg/L	10	Standard
Fe	54	556.2	7.6	**0.2841**	0.037	12.9	mg/L	179	Standard
Fe	57	410.0	16.9	**0.1734**	0.212	122.1	mg/L	335	Standard
> **Sc-1**	45	68191.6	3.4				mg/L	61029	Standard
Cl	35	26725.6	1.1				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	35244.3	3.6				ug/L	1547	Standard
P	31	35015.4	1.7				ug/L	15286	Standard
S	34	2091.8	5.9				ug/L	1572	Standard
Sr	88	225.0	20.0				ug/L	73	Standard
C	12	103.3	39.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	93.3	12.4				mg/L	3	Standard
Dy	164	1034.1	12.7				mg/L	6	Standard
Ho-1	165	763.4	3.9				mg/L	10	Standard
Er	166	680.0	15.4				mg/L	13	Standard
I	127	80476.5	1.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.292	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.426	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606102604
Report Date/Time: Wednesday, June 29, 2016 14:56:48
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.933
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		99.491
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Ba 135 Upper, S, EEE	Ba	135	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102606
Sample Date/Time: Wednesday, June 29, 2016 14:57:42
Number of Replicates: 3
Autosampler Position: 334
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93356.7	0.9				ug/L	74699	Standard
	Be	9	136.7	43.6	**0.1318**	0.054	40.9	ug/L	8	Standard
	Al	27	4727947.0	2.8	**25.7324**	0.870	3.4	ug/L	1493	Standard
	Sc	45	71265.6	1.4				ug/L	61029	Standard
	Ti	47	7167.7	1.1	**18.5395**	0.116	0.6	ug/L	35	Standard
	V	51	33631.6	0.4	**4.1963**	0.093	2.2	ug/L	1435	Standard
	Cr	52	43684.3	3.9	**4.7878**	0.170	3.6	ug/L	7635	Standard
	Cr	53	18915.1	0.4	**19.8540**	0.330	1.7	ug/L	363	Standard
	Mn	55	527300.2	1.1	**74.5358**	0.480	0.6	ug/L	747	Standard
	Co	59	47125.2	1.5	**6.9996**	0.016	0.2	ug/L	234	Standard
	Ni	60	57578.9	1.0	**30.2025**	0.274	0.9	ug/L	215	Standard
	Cu	65	3787.5	1.8	**1.9028**	0.007	0.4	ug/L	104	Standard
	Zn	66	29686.7	1.5	**26.3448**	0.262	1.0	ug/L	363	Standard
>	**Ge**	72	502153.4	1.7				ug/L	495142	Standard
	As	75	1552.5	5.8	**1.4257**	0.058	4.0	ug/L	-50	Standard
	Se	82	328.6	4.9	**2.8136**	0.101	3.6	ug/L	14	Standard
	Se-1	77	866.0	3.6	**11.0672**	0.235	2.1	ug/L	42	Standard
>	**Ga**	71	2006.8	3.5				mg/L	42	Standard
	Rb	85	28431.7	2.8				ug/L	17	Standard
	Y	89	467809.8	1.5				ug/L	417098	Standard
>	**Rh**	103	28.3	36.7				ug/L	13	Standard
	Mo	98	2788.7	2.4	**1.0189**	0.018	1.8	ug/L	11	Standard
	Ag	107	187.0	8.5	**0.0152**	0.002	14.7	ug/L	68	Standard
	Cd	111	184.5	8.6	**0.0785**	0.007	8.5	mg/L	2	Standard
	Cd	114	502.9	12.6	**0.0940**	0.013	13.9	ug/L	12	Standard
>	**In**	115	599816.0	1.2				ug/L	587441	Standard
	Sn	118	1385.1	2.5	**0.1375**	0.003	2.2	ug/L	497	Standard
	Sb	123	315.5	7.5	**0.0620**	0.005	7.5	ug/L	78	Standard
	Ba	135	37805.8	2.1	**18.0728**	0.162	0.9	ug/L	24	Standard
	Ce	140	83749.2	1.1				ug/L	25	Standard
>	**Tb**	159	971524.4	1.2				ug/L	939701	Standard
	Ho	165	1393.4	2.1				ug/L	10	Standard
	Tl	203	289.7	7.1	**0.0339**	0.002	6.5	ug/L	4	Standard
	Tl	205	258.3	11.0	**0.0368**	0.004	11.7	ug/L	2	Standard
	Pb	206	3853.2	1.2	**0.6604**	0.004	0.6	ug/L	240	Standard
	Pb	207	3010.0	1.5	**0.5732**	0.009	1.6	ug/L	198	Standard
	Pb	208	13337.3	2.3	**0.6473**	0.013	1.9	ug/L	883	Standard
	U	238	859.0	4.4	**0.1221**	0.005	4.0	ug/L	3	Standard
>	**Bi**	209	516635.2	0.9				ug/L	523984	Standard

Sample ID: L1606102606
Report Date/Time: Wednesday, June 29, 2016 14:59:59
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.3563**	28.321	173.2	mg/L	0	Standard
Mg	24	103702.1	2.0	**114.5988**	3.552	3.1	mg/L	37	Standard
K	39	251.7	12.0	**0.3627**	0.062	17.2	mg/L	38	Standard
Ca	43	61.7	26.1	**0.8208**	0.922	112.4	mg/L	10	Standard
Fe	54	1412.4	2.3	**0.8871**	0.015	1.7	mg/L	179	Standard
Fe	57	645.0	2.8	**0.8030**	0.026	3.3	mg/L	335	Standard
Sc-1	45	71265.6	1.4				mg/L	61029	Standard
Cl	35	28407.3	0.4				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	49253.3	3.6				ug/L	1547	Standard
P	31	41409.9	2.5				ug/L	15286	Standard
S	34	2068.5	4.4				ug/L	1572	Standard
Sr	88	155.0	22.6				ug/L	73	Standard
C	12	133.3	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	57.3				mg/L	3	Standard
Dy	164	2353.9	2.8				mg/L	6	Standard
Ho-1	165	1393.4	2.1				mg/L	10	Standard
Er	166	1453.4	9.2				mg/L	13	Standard
I	127	3067238.7	11.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		124.978	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		101.416	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606102606
Report Date/Time: Wednesday, June 29, 2016 14:59:59
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	102.107	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	98.597	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Sample ID: L1606102606
Report Date/Time: Wednesday, June 29, 2016 14:59:59
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606102607
Sample Date/Time: Wednesday, June 29, 2016 15:00:54
Number of Replicates: 3
Autosampler Position: 335
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86025.2	1.8				ug/L	74699	Standard
	Be	9	83.3	29.6	**0.0901**	0.026	28.3	ug/L	8	Standard
	Al	27	3430436.9	1.2	**20.2582**	0.272	1.3	ug/L	1493	Standard
	Sc	45	67818.2	2.7				ug/L	61029	Standard
	Ti	47	2650.2	2.1	**6.1960**	0.095	1.5	ug/L	35	Standard
	V	51	9706.1	1.4	**0.9634**	0.025	2.6	ug/L	1435	Standard
	Cr	52	14237.0	2.0	**0.7086**	0.035	4.9	ug/L	7635	Standard
	Cr	53	2521.9	11.8	**2.1295**	0.275	12.9	ug/L	363	Standard
	Mn	55	50752.2	3.1	**6.4246**	0.167	2.6	ug/L	747	Standard
	Co	59	8541.7	4.1	**1.1290**	0.040	3.6	ug/L	234	Standard
	Ni	60	2905.6	5.8	**1.2753**	0.074	5.8	ug/L	215	Standard
	Cu	65	1384.7	3.1	**0.5672**	0.017	3.0	ug/L	104	Standard
	Zn	66	5145.2	1.6	**3.7928**	0.039	1.0	ug/L	363	Standard
>	**Ge**	72	549096.7	0.7				ug/L	495142	Standard
	As	75	307.3	24.2	**0.3069**	0.059	19.1	ug/L	-50	Standard
	Se	82	131.2	8.7	**0.9395**	0.088	9.3	ug/L	14	Standard
	Se-1	77	177.7	11.5	**1.5705**	0.241	15.4	ug/L	42	Standard
>	**Ga**	71	765.0	18.4				mg/L	42	Standard
	Rb	85	11574.4	1.2				ug/L	17	Standard
	Y	89	476523.2	1.2				ug/L	417098	Standard
>	**Rh**	103	40.0	54.5				ug/L	13	Standard
	Mo	98	552.0	3.2	**0.1876**	0.008	4.1	ug/L	11	Standard
	Ag	107	103.3	5.6	**0.0019**	0.001	51.8	ug/L	68	Standard
	Cd	111	131.7	9.5	**0.0509**	0.006	12.0	mg/L	2	Standard
	Cd	114	303.7	9.9	**0.0531**	0.005	10.1	ug/L	12	Standard
>	**In**	115	629575.3	1.6				ug/L	587441	Standard
	Sn	118	1851.8	1.5	**0.2011**	0.002	0.9	ug/L	497	Standard
	Sb	123	475.2	2.5	**0.0908**	0.004	4.2	ug/L	78	Standard
	Ba	135	119108.2	2.3	**54.2912**	1.481	2.7	ug/L	24	Standard
	Ce	140	65755.7	1.8				ug/L	25	Standard
>	**Tb**	159	1004571.7	2.5				ug/L	939701	Standard
	Ho	165	1471.7	8.7				ug/L	10	Standard
	Tl	203	123.3	5.2	**0.0121**	0.001	4.7	ug/L	4	Standard
	Tl	205	98.3	10.6	**0.0129**	0.002	11.8	ug/L	2	Standard
	Pb	206	2854.9	1.5	**0.4229**	0.014	3.4	ug/L	240	Standard
	Pb	207	2325.8	2.5	**0.3831**	0.018	4.6	ug/L	198	Standard
	Pb	208	9906.6	0.9	**0.4144**	0.011	2.7	ug/L	883	Standard
	U	238	660.3	1.2	**0.0843**	0.002	2.3	ug/L	3	Standard
>	**Bi**	209	566874.3	1.4				ug/L	523984	Standard

Sample ID: L1606102607
Report Date/Time: Wednesday, June 29, 2016 15:03:11
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	9059.4	5.1	**10.4668**	0.674	6.4	mg/L	37	Standard
K	39	175.0	7.6	**0.2337**	0.017	7.4	mg/L	38	Standard
Ca	43	76.7	10.0	**1.8315**	0.359	19.6	mg/L	10	Standard
Fe	54	778.8	6.8	**0.4564**	0.046	10.1	mg/L	179	Standard
Fe	57	450.0	7.8	**0.3039**	0.129	42.4	mg/L	335	Standard
> **Sc-1**	45	67818.2	2.7				mg/L	61029	Standard
Cl	35	27265.2	1.0				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	4797.4	2.1				ug/L	1547	Standard
P	31	22328.1	2.9				ug/L	15286	Standard
S	34	1868.4	11.0				ug/L	1572	Standard
Sr	88	170.0	23.0				ug/L	73	Standard
C	12	116.7	4.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	2276.8	3.2				mg/L	6	Standard
Ho-1	165	1471.7	8.7				mg/L	10	Standard
Er	166	1323.4	4.9				mg/L	13	Standard
I	127	142556.3	31.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		115.163	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.897	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606102607
Report Date/Time: Wednesday, June 29, 2016 15:03:11
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		107.172
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.185
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 15:04:07
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	83434.4	4.1				ug/L	74699	Standard
	Be	9	52151.3	1.6	**53.0799**	1.438	2.7	ug/L	8	Standard
	Al	27	8001829.5	1.2	**48.7762**	1.563	3.2	ug/L	1493	Standard
	Sc	45	63472.7	2.1				ug/L	61029	Standard
	Ti	47	41544.5	0.5	**100.8598**	1.507	1.5	ug/L	35	Standard
	V	51	430823.0	0.6	**52.3478**	0.567	1.1	ug/L	1435	Standard
	Cr	52	414183.8	1.0	**50.4520**	0.020	0.0	ug/L	7635	Standard
	Cr	53	51435.5	2.1	**50.9262**	0.597	1.2	ug/L	363	Standard
	Mn	55	402509.0	2.2	**53.1108**	1.647	3.1	ug/L	747	Standard
	Co	59	395807.1	0.7	**55.1753**	0.902	1.6	ug/L	234	Standard
	Ni	60	102515.7	1.1	**50.3079**	0.215	0.4	ug/L	215	Standard
	Cu	65	102297.5	1.1	**50.5232**	0.685	1.4	ug/L	104	Standard
	Zn	66	61288.7	1.4	**51.2272**	0.860	1.7	ug/L	363	Standard
>	Ge	72	537577.7	1.0				ug/L	495142	Standard
	As	75	61234.0	0.5	**50.4168**	0.738	1.5	ug/L	-50	Standard
	Se	82	6132.0	1.1	**51.3300**	1.089	2.1	ug/L	14	Standard
	Se-1	77	4299.6	0.9	**53.6055**	0.977	1.8	ug/L	42	Standard
>	Ga	71	78.3	42.5				mg/L	42	Standard
	Rb	85	866.7	7.4				ug/L	17	Standard
	Y	89	460058.1	1.7				ug/L	417098	Standard
>	Rh	103	33.3	31.2				ug/L	13	Standard
	Mo	98	295506.5	0.5	**103.9241**	2.114	2.0	ug/L	11	Standard
	Ag	107	348188.0	0.5	**50.0657**	0.748	1.5	ug/L	68	Standard
	Cd	111	111062.0	1.3	**49.7777**	0.346	0.7	mg/L	2	Standard
	Cd	114	271614.5	0.9	**49.6432**	0.990	2.0	ug/L	12	Standard
>	In	115	626816.0	1.9				ug/L	587441	Standard
	Sn	118	313548.1	1.3	**50.2246**	1.384	2.8	ug/L	497	Standard
	Sb	123	246409.0	0.1	**49.4692**	0.867	1.8	ug/L	78	Standard
	Ba	135	105775.6	0.8	**48.4328**	1.255	2.6	ug/L	24	Standard
	Ce	140	66.7	48.8				ug/L	25	Standard
>	Tb	159	996778.4	1.8				ug/L	939701	Standard
	Ho	165	16.7	34.6				ug/L	10	Standard
	Tl	203	442598.6	0.7	**50.9461**	0.764	1.5	ug/L	4	Standard
	Tl	205	402531.2	1.7	**53.3816**	0.705	1.3	ug/L	2	Standard
	Pb	206	287928.0	0.2	**51.0713**	0.393	0.8	ug/L	240	Standard
	Pb	207	257921.1	0.2	**51.4941**	0.566	1.1	ug/L	198	Standard
	Pb	208	1040092.9	0.4	**52.5571**	0.545	1.0	ug/L	883	Standard
	U	238	381264.5	1.2	**52.5495**	0.479	0.9	ug/L	3	Standard
>	Bi	209	552262.9	0.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:06:24
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4305.6	5.5	**5.2771**	0.213	4.0	mg/L	37	Standard
K	39	2275.2	2.4	**4.7239**	0.217	4.6	mg/L	38	Standard
Ca	43	126.7	42.6	**5.1579**	3.228	62.6	mg/L	10	Standard
Fe	54	6415.2	4.5	**5.0956**	0.328	6.4	mg/L	179	Standard
Fe	57	1851.8	4.1	**4.9827**	0.373	7.5	mg/L	335	Standard
> **Sc-1**	45	63472.7	2.1				mg/L	61029	Standard
Cl	35	28119.4	1.3				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2310.2	13.5				ug/L	1547	Standard
P	31	19991.5	1.1				ug/L	15286	Standard
S	34	2273.5	5.1				ug/L	1572	Standard
Sr	88	71.7	10.7				ug/L	73	Standard
C	12	66.7	8.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	8.6	116.8				mg/L	6	Standard
Ho-1	165	16.7	34.6				mg/L	10	Standard
Er	166	30.0	0.0				mg/L	13	Standard
I	127	25116.6	23.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	106.160		
Al	27	97.552		
Sc	45			
Ti	47	100.860		
V	51	104.696		
Cr	52	100.904		
Cr	53			
Mn	55	106.222		
Co	59	110.351		
Ni	60	100.616		
Cu	65	101.046		
Zn	66	102.454		
> Ge	72		108.570	
As	75	100.834		
Se	82	102.660		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:06:24
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	103.924		
Ag	107	100.131		
Cd	111	99.555		
Cd	114			
In	115		106.703	
Sn	118	100.449		
Sb	123	98.938		
Ba	135	96.866		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.892		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.114		
U	238	105.099		
Bi	209		105.397	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:06:24
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 15:07:18
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	85012.9	2.0				ug/L	74699	Standard
	Be	9	28.3	27.0	**0.0359**	0.008	21.1	ug/L	8	Standard
	Al	27	3885.6	39.7	**0.0100**	0.010	97.9	ug/L	1493	Standard
	Sc	45	66216.1	1.0				ug/L	61029	Standard
	Ti	47	42.3	11.2	**-0.0066**	0.011	167.9	ug/L	35	Standard
	V	51	1610.8	4.5	**0.0001**	0.010	18188.0	ug/L	1435	Standard
	Cr	52	7733.0	4.2	**-0.0667**	0.042	62.7	ug/L	7635	Standard
	Cr	53	656.7	7.6	**0.3218**	0.049	15.4	ug/L	363	Standard
	Mn	55	934.4	2.8	**-0.0250**	0.004	15.9	ug/L	747	Standard
	Co	59	270.3	16.3	**0.0003**	0.006	2278.1	ug/L	234	Standard
	Ni	60	353.0	46.0	**0.0492**	0.080	162.9	ug/L	215	Standard
	Cu	65	160.7	5.6	**-0.0245**	0.004	18.2	ug/L	104	Standard
	Zn	66	427.3	1.4	**-0.0957**	0.006	6.2	ug/L	363	Standard
>	Ge	72	539852.0	0.5				ug/L	495142	Standard
	As	75	-35.2	72.3	**0.0308**	0.021	67.1	ug/L	-50	Standard
	Se	82	23.4	13.4	**0.0577**	0.027	47.1	ug/L	14	Standard
	Se-1	77	65.0	15.6	**0.1944**	0.131	67.4	ug/L	42	Standard
>	Ga	71	48.3	15.8				mg/L	42	Standard
	Rb	85	13.3	21.7				ug/L	17	Standard
	Y	89	460334.1	1.4				ug/L	417098	Standard
>	Rh	103	15.0	57.7				ug/L	13	Standard
	Mo	98	325.6	30.7	**0.1063**	0.035	33.0	ug/L	11	Standard
	Ag	107	121.0	14.9	**0.0041**	0.003	62.9	ug/L	68	Standard
	Cd	111	13.2	31.2	**-0.0021**	0.002	85.0	mg/L	2	Standard
	Cd	114	51.9	43.2	**0.0072**	0.004	56.5	ug/L	12	Standard
>	In	115	641477.2	0.6				ug/L	587441	Standard
	Sn	118	860.0	23.5	**0.0400**	0.031	77.0	ug/L	497	Standard
	Sb	123	396.3	20.9	**0.0735**	0.016	21.4	ug/L	78	Standard
	Ba	135	45.3	20.4	**0.0043**	0.004	99.1	ug/L	24	Standard
	Ce	140	36.7	15.7				ug/L	25	Standard
>	Tb	159	1015779.3	0.6				ug/L	939701	Standard
	Ho	165	5.0	100.0				ug/L	10	Standard
	Tl	203	76.0	35.5	**0.0066**	0.003	43.8	ug/L	4	Standard
	Tl	205	63.3	18.2	**0.0082**	0.001	17.5	ug/L	2	Standard
	Pb	206	309.7	4.7	**-0.0185**	0.002	12.2	ug/L	240	Standard
	Pb	207	263.0	6.6	**-0.0197**	0.003	14.7	ug/L	198	Standard
	Pb	208	1081.0	8.8	**-0.0217**	0.004	19.1	ug/L	883	Standard
	U	238	55.3	39.0	**0.0029**	0.003	97.3	ug/L	3	Standard
>	Bi	209	576716.4	0.8				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:09:35
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.8612**	30.928	173.2	mg/L	0	Standard
Mg	24	48.3	39.2	**-0.0075**	0.023	309.1	mg/L	37	Standard
K	39	65.0	38.5	**0.0181**	0.050	275.9	mg/L	38	Standard
Ca	43	31.7	9.1	**-0.6795**	0.172	25.4	mg/L	10	Standard
Fe	54	165.7	15.0	**-0.0089**	0.019	210.6	mg/L	179	Standard
Fe	57	258.3	12.9	**-0.2654**	0.099	37.5	mg/L	335	Standard
Sc-1	45	66216.1	1.0				mg/L	61029	Standard
Cl	35	27856.9	0.3				ug/L	20195	Standard
Kr	83	4.0	43.3				ug/L	4	Standard
Br	81	1983.5	3.1				ug/L	1547	Standard
P	31	20308.6	3.1				ug/L	15286	Standard
S	34	2271.8	3.7				ug/L	1572	Standard
Sr	88	51.7	22.3				ug/L	73	Standard
C	12	73.3	34.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	20.0	86.6				mg/L	6	Standard
Ho-1	165	5.0	100.0				mg/L	10	Standard
Er	166	0.0					mg/L	13	Standard
I	127	13137.4	11.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		109.030	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:09:35
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.199
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.064
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:09:35
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBS M5 WG573099-02
Sample Date/Time: Wednesday, June 29, 2016 15:10:32
Number of Replicates: 3
Autosampler Position: 336
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82837.5	1.4				ug/L	74699	Standard
	Be	9	20.0	75.0	**0.0280**	0.015	53.9	ug/L	8	Standard
	Al	27	15988.5	12.5	**0.0848**	0.014	16.0	ug/L	1493	Standard
	Sc	45	64671.1	2.4				ug/L	61029	Standard
	Ti	47	45.0	20.0	**-0.0015**	0.023	1562.0	ug/L	35	Standard
	V	51	1647.0	1.8	**0.0017**	0.004	236.7	ug/L	1435	Standard
	Cr	52	11218.5	4.3	**0.3456**	0.052	15.0	ug/L	7635	Standard
	Cr	53	1063.4	5.0	**0.7116**	0.068	9.6	ug/L	363	Standard
	Mn	55	1756.1	5.1	**0.0801**	0.011	14.2	ug/L	747	Standard
	Co	59	312.3	6.3	**0.0055**	0.003	48.1	ug/L	234	Standard
	Ni	60	1377.4	15.3	**0.5415**	0.099	18.4	ug/L	215	Standard
	Cu	65	261.7	8.1	**0.0235**	0.011	45.2	ug/L	104	Standard
	Zn	66	1613.1	3.7	**0.8814**	0.040	4.6	ug/L	363	Standard
>	**Ge**	72	547319.5	1.5				ug/L	495142	Standard
	As	75	-42.5	100.6	**0.0255**	0.034	133.0	ug/L	-50	Standard
	Se	82	18.0	18.7	**0.0100**	0.027	272.2	ug/L	14	Standard
	Se-1	77	65.0	16.6	**0.1841**	0.145	78.6	ug/L	42	Standard
>	**Ga**	71	66.7	22.9				mg/L	42	Standard
	Rb	85	36.7	56.8				ug/L	17	Standard
	Y	89	450480.7	0.9				ug/L	417098	Standard
>	**Rh**	103	26.7	10.8				ug/L	13	Standard
	Mo	98	139.4	1.7	**0.0435**	0.001	2.2	ug/L	11	Standard
	Ag	107	102.7	18.1	**0.0019**	0.003	140.0	ug/L	68	Standard
	Cd	111	13.1	71.9	**-0.0020**	0.004	209.6	mg/L	2	Standard
	Cd	114	24.5	136.0	**0.0024**	0.006	256.9	ug/L	12	Standard
>	**In**	115	624818.4	0.5				ug/L	587441	Standard
	Sn	118	765.0	2.8	**0.0284**	0.003	12.0	ug/L	497	Standard
	Sb	123	189.1	20.4	**0.0339**	0.008	23.4	ug/L	78	Standard
	Ba	135	83.0	7.9	**0.0221**	0.003	14.0	ug/L	24	Standard
	Ce	140	301.7	18.8				ug/L	25	Standard
>	**Tb**	159	992078.8	0.6				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	40.0	57.5	**0.0027**	0.003	94.2	ug/L	4	Standard
	Tl	205	43.3	29.0	**0.0058**	0.002	28.5	ug/L	2	Standard
	Pb	206	319.7	11.1	**-0.0160**	0.006	36.6	ug/L	240	Standard
	Pb	207	262.0	4.6	**-0.0191**	0.002	12.7	ug/L	198	Standard
	Pb	208	1094.7	2.8	**-0.0202**	0.002	8.1	ug/L	883	Standard
	U	238	22.0	58.0	**-0.0015**	0.002	114.9	ug/L	3	Standard
>	**Bi**	209	568214.1	0.6				ug/L	523984	Standard

Sample ID: PBS M5 WG573099-02
Report Date/Time: Wednesday, June 29, 2016 15:12:48
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	43.3	17.6	**-0.0124**	0.008	67.7	mg/L	37	Standard
K	39	35.0	51.5	**-0.0416**	0.036	86.5	mg/L	38	Standard
Ca	43	30.0	28.9	**-0.7286**	0.543	74.5	mg/L	10	Standard
Fe	54	174.1	18.8	**0.0014**	0.029	2030.6	mg/L	179	Standard
Fe	57	266.7	22.0	**-0.2161**	0.204	94.4	mg/L	335	Standard
> **Sc-1**	45	64671.1	2.4				mg/L	61029	Standard
Cl	35	28617.0	1.0				ug/L	20195	Standard
Kr	83	2.0	100.0				ug/L	4	Standard
Br	81	1953.5	6.4				ug/L	1547	Standard
P	31	19132.1	3.4				ug/L	15286	Standard
S	34	2278.5	6.0				ug/L	1572	Standard
Sr	88	68.3	29.6				ug/L	73	Standard
C	12	83.3	13.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.5	39.0				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	23.3	49.5				mg/L	13	Standard
I	127	8569.1	14.4				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		110.895	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		110.538	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBS M5 WG573099-02
Report Date/Time: Wednesday, June 29, 2016 15:12:48
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.363
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.441
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBS M5 WG573099-02
Report Date/Time: Wednesday, June 29, 2016 15:12:48
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSS M5 WG573099-03
Sample Date/Time: Wednesday, June 29, 2016 15:13:43
Number of Replicates: 3
Autosampler Position: 337
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83551.5	2.3				ug/L	74699	Standard
	Be	9	25751.9	2.9	**26.1593**	0.514	2.0	ug/L	8	Standard
	Al	27	17496.8	6.9	**0.0930**	0.006	6.2	ug/L	1493	Standard
	Sc	45	66115.6	1.8				ug/L	61029	Standard
	Ti	47	52.7	14.3	**0.0175**	0.016	92.6	ug/L	35	Standard
	V	51	220707.2	1.1	**26.4464**	0.390	1.5	ug/L	1435	Standard
	Cr	52	221818.0	0.7	**26.2660**	0.546	2.1	ug/L	7635	Standard
	Cr	53	27359.7	0.9	**26.6591**	0.406	1.5	ug/L	363	Standard
	Mn	55	204818.9	1.5	**26.6717**	0.451	1.7	ug/L	747	Standard
	Co	59	196776.0	1.4	**27.1279**	0.304	1.1	ug/L	234	Standard
	Ni	60	55651.4	0.4	**26.9765**	0.492	1.8	ug/L	215	Standard
	Cu	65	54836.4	0.8	**26.7594**	0.556	2.1	ug/L	104	Standard
	Zn	66	32934.1	1.1	**27.0387**	0.809	3.0	ug/L	363	Standard
>	**Ge**	72	543210.6	2.2				ug/L	495142	Standard
	As	75	31720.5	1.5	**25.8765**	0.351	1.4	ug/L	-50	Standard
	Se	82	3103.9	1.9	**25.6428**	0.383	1.5	ug/L	14	Standard
	Se-1	77	2231.2	2.7	**27.2211**	0.278	1.0	ug/L	42	Standard
>	**Ga**	71	55.0	47.2				mg/L	42	Standard
	Rb	85	35.0	42.9				ug/L	17	Standard
	Y	89	467716.3	0.8				ug/L	417098	Standard
>	**Rh**	103	13.3	43.3				ug/L	13	Standard
	Mo	98	194.8	10.9	**0.0603**	0.008	12.8	ug/L	11	Standard
	Ag	107	10349.2	1.8	**1.4201**	0.038	2.7	ug/L	68	Standard
	Cd	111	56717.0	1.5	**24.4699**	0.504	2.1	mg/L	2	Standard
	Cd	114	136634.8	1.7	**24.0382**	0.506	2.1	ug/L	12	Standard
>	**In**	115	651063.9	0.8				ug/L	587441	Standard
	Sn	118	761.7	5.5	**0.0229**	0.007	28.6	ug/L	497	Standard
	Sb	123	128047.0	1.0	**24.7441**	0.436	1.8	ug/L	78	Standard
	Ba	135	54350.3	0.9	**23.9455**	0.395	1.6	ug/L	24	Standard
	Ce	140	61.7	26.1				ug/L	25	Standard
>	**Tb**	159	1033159.0	0.5				ug/L	939701	Standard
	Ho	165	16.7	17.3				ug/L	10	Standard
	Tl	203	227266.1	0.5	**24.7175**	0.218	0.9	ug/L	4	Standard
	Tl	205	203485.9	1.6	**25.5015**	0.511	2.0	ug/L	2	Standard
	Pb	206	152382.5	0.6	**25.5050**	0.255	1.0	ug/L	240	Standard
	Pb	207	130131.8	0.7	**24.5130**	0.243	1.0	ug/L	198	Standard
	Pb	208	530327.1	0.9	**25.2835**	0.262	1.0	ug/L	883	Standard
	U	238	189200.1	1.2	**24.6397**	0.334	1.4	ug/L	3	Standard
>	**Bi**	209	584433.1	0.4				ug/L	523984	Standard

Sample ID: LCSS M5 WG573099-03
Report Date/Time: Wednesday, June 29, 2016 15:16:00
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	41.7	36.7	**-0.0156**	0.017	112.1	mg/L	37	Standard
K	39	38.3	52.7	**-0.0354**	0.042	119.1	mg/L	38	Standard
Ca	43	25.0	52.9	**-1.0705**	0.758	70.8	mg/L	10	Standard
Fe	54	141.4	21.3	**-0.0274**	0.025	92.5	mg/L	179	Standard
Fe	57	241.7	15.7	**-0.3151**	0.126	40.1	mg/L	335	Standard
> **Sc-1**	45	66115.6	1.8				mg/L	61029	Standard
Cl	35	28385.2	0.6				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	2056.8	1.5				ug/L	1547	Standard
P	31	18911.8	2.1				ug/L	15286	Standard
S	34	2256.8	4.9				ug/L	1572	Standard
Sr	88	70.0	39.8				ug/L	73	Standard
C	12	60.0	66.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	9.5	5.0				mg/L	6	Standard
Ho-1	165	16.7	17.3				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	6019.5	14.1				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		111.851	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		109.708	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: LCSS M5 WG573099-03
Report Date/Time: Wednesday, June 29, 2016 15:16:00
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.831
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.536
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSS M5 WG573099-03
Report Date/Time: Wednesday, June 29, 2016 15:16:00
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088803 WG573099-01
Sample Date/Time: Wednesday, June 29, 2016 15:16:55
Number of Replicates: 3
Autosampler Position: 338
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	87864.2	2.8				ug/L	74699	Standard
	Be	9	1871.8	7.3	**1.8142**	0.100	5.5	ug/L	8	Standard
	Al	27	17464438.5	2.2	**101.0382**	1.686	1.7	ug/L	1493	Standard
	Sc	45	88916.8	1.0				ug/L	61029	Standard
	Ti	47	8105.2	1.6	**20.2503**	0.364	1.8	ug/L	35	Standard
	V	51	266665.9	0.8	**33.4185**	0.188	0.6	ug/L	1435	Standard
	Cr	52	308197.4	0.8	**38.5684**	0.415	1.1	ug/L	7635	Standard
	Cr	53	37658.4	2.2	**38.4602**	0.742	1.9	ug/L	363	Standard
	Mn	55	19789542.3	0.8	**2705.9402**	31.717	1.2	ug/L	747	Standard
	Co	59	276278.5	0.6	**39.7944**	0.401	1.0	ug/L	234	Standard
	Ni	60	160372.4	1.4	**81.4241**	1.275	1.6	ug/L	215	Standard
	Cu	65	107769.0	1.4	**55.0229**	0.902	1.6	ug/L	104	Standard
	Zn	66	213527.8	0.7	**185.6475**	1.406	0.8	ug/L	363	Standard
>	**Ge**	72	520092.2	0.4				ug/L	495142	Standard
	As	75	6170.4	1.0	**5.3042**	0.070	1.3	ug/L	-50	Standard
	Se	82	90.8	2.7	**0.6494**	0.023	3.5	ug/L	14	Standard
	Se-1	77	157.7	4.9	**1.4330**	0.106	7.4	ug/L	42	Standard
>	**Ga**	71	47343.9	1.1				mg/L	42	Standard
	Rb	85	380728.2	1.5				ug/L	17	Standard
	Y	89	1106618.3	0.2				ug/L	417098	Standard
>	**Rh**	103	63.3	24.1				ug/L	13	Standard
	Mo	98	980.4	3.2	**0.3366**	0.010	3.0	ug/L	11	Standard
	Ag	107	955.0	6.8	**0.1236**	0.011	9.1	ug/L	68	Standard
	Cd	111	2225.7	2.0	**0.9829**	0.017	1.8	mg/L	2	Standard
	Cd	114	5479.9	1.8	**0.9924**	0.001	0.1	ug/L	12	Standard
>	**In**	115	631161.9	1.7				ug/L	587441	Standard
	Sn	118	413.3	14.8	**-0.0289**	0.009	31.0	ug/L	497	Standard
	Sb	123	370.7	34.4	**0.0695**	0.025	35.4	ug/L	78	Standard
	Ba	135	385428.4	1.0	**175.2729**	2.367	1.4	ug/L	24	Standard
	Ce	140	2007459.4	3.6				ug/L	25	Standard
>	**Tb**	159	1048778.2	1.5				ug/L	939701	Standard
	Ho	165	53168.1	1.9				ug/L	10	Standard
	Tl	203	2332.8	1.5	**0.2668**	0.007	2.6	ug/L	4	Standard
	Tl	205	2056.8	1.1	**0.2730**	0.001	0.5	ug/L	2	Standard
	Pb	206	102474.4	0.3	**18.1326**	0.189	1.0	ug/L	240	Standard
	Pb	207	82544.6	1.5	**16.4326**	0.141	0.9	ug/L	198	Standard
	Pb	208	353310.3	0.4	**17.8060**	0.180	1.0	ug/L	883	Standard
	U	238	9355.9	0.7	**1.2854**	0.017	1.3	ug/L	3	Standard
>	**Bi**	209	552227.6	1.4				ug/L	523984	Standard

Sample ID: L1606088803 WG573099-01
Report Date/Time: Wednesday, June 29, 2016 15:19:11
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**13.3403**	23.097	173.1	mg/L	0	Standard
Mg	24	320.0	5.6	**0.2185**	0.016	7.5	mg/L	37	Standard
K	39	1080.0	10.5	**1.5233**	0.156	10.2	mg/L	38	Standard
Ca	43	916.7	12.0	**37.2112**	4.625	12.4	mg/L	10	Standard
Fe	54	75887.5	2.1	**44.0256**	1.050	2.4	mg/L	179	Standard
Fe	57	19502.5	1.2	**44.4256**	0.652	1.5	mg/L	335	Standard
> **Sc-1**	45	88916.8	1.0				mg/L	61029	Standard
Cl	35	27513.6	0.5				ug/L	20195	Standard
Kr	83	0.3	173.2				ug/L	4	Standard
Br	81	3467.1	2.9				ug/L	1547	Standard
P	31	21865.8	2.4				ug/L	15286	Standard
S	34	2631.9	4.0				ug/L	1572	Standard
Sr	88	220.0	6.0				ug/L	73	Standard
C	12	66.7	37.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	36.7	15.7				mg/L	3	Standard
Dy	164	86599.4	2.5				mg/L	6	Standard
Ho-1	165	53168.1	1.9				mg/L	10	Standard
Er	166	47260.3	0.7				mg/L	13	Standard
I	127	6809.9	9.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		117.625	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		105.039	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606088803 WG573099-01
Report Date/Time: Wednesday, June 29, 2016 15:19:11
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		107.443
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.390
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606088803 WG573099-01
Report Date/Time: Wednesday, June 29, 2016 15:19:11
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088803S WG573099-04
Sample Date/Time: Wednesday, June 29, 2016 15:20:06
Number of Replicates: 3
Autosampler Position: 339
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88784.5	2.9				ug/L	74699	Standard
	Be	9	27508.3	0.4	**26.3080**	0.631	2.4	ug/L	8	Standard
	Al	27	20794150.5	3.6	**119.0173**	0.902	0.8	ug/L	1493	Standard
	Sc	45	96992.9	2.3				ug/L	61029	Standard
	Ti	47	11606.4	2.4	**28.2553**	0.997	3.5	ug/L	35	Standard
	V	51	594819.7	3.2	**72.7593**	3.186	4.4	ug/L	1435	Standard
	Cr	52	602098.1	1.8	**74.2282**	2.325	3.1	ug/L	7635	Standard
	Cr	53	74295.1	1.5	**74.1206**	1.783	2.4	ug/L	363	Standard
	Mn	55	28090211.5	1.2	**3736.5671**	85.696	2.3	ug/L	747	Standard
	Co	59	552502.8	1.4	**77.4533**	1.929	2.5	ug/L	234	Standard
	Ni	60	270132.5	1.6	**133.5117**	3.898	2.9	ug/L	215	Standard
	Cu	65	201432.2	2.2	**100.1412**	3.335	3.3	ug/L	104	Standard
	Zn	66	320182.0	2.8	**271.0655**	11.001	4.1	ug/L	363	Standard
>	**Ge**	72	534719.5	1.4				ug/L	495142	Standard
	As	75	41934.7	2.3	**34.7370**	1.197	3.4	ug/L	-50	Standard
	Se	82	2480.2	1.5	**20.7934**	0.607	2.9	ug/L	14	Standard
	Se-1	77	1799.1	2.3	**22.1951**	0.836	3.8	ug/L	42	Standard
>	**Ga**	71	63861.1	4.0				mg/L	42	Standard
	Rb	85	554637.9	1.8				ug/L	17	Standard
	Y	89	1250941.0	3.4				ug/L	417098	Standard
>	**Rh**	103	88.3	8.6				ug/L	13	Standard
	Mo	98	1454.1	4.8	**0.5217**	0.033	6.4	ug/L	11	Standard
	Ag	107	4814.4	4.1	**0.7006**	0.018	2.6	ug/L	68	Standard
	Cd	111	56224.0	1.3	**25.9728**	0.111	0.4	mg/L	2	Standard
	Cd	114	138439.7	3.6	**26.0717**	0.493	1.9	ug/L	12	Standard
>	**In**	115	608045.3	1.7				ug/L	587441	Standard
	Sn	118	518.3	2.0	**-0.0090**	0.001	16.4	ug/L	497	Standard
	Sb	123	2325.7	1.7	**0.4771**	0.008	1.7	ug/L	78	Standard
	Ba	135	587827.7	1.5	**277.4602**	0.795	0.3	ug/L	24	Standard
	Ce	140	2927642.8	0.6				ug/L	25	Standard
>	**Tb**	159	1033681.4	2.7				ug/L	939701	Standard
	Ho	165	68191.4	0.6				ug/L	10	Standard
	Tl	203	212892.3	0.8	**25.4932**	0.083	0.3	ug/L	4	Standard
	Tl	205	191048.7	1.2	**26.3598**	0.091	0.3	ug/L	2	Standard
	Pb	206	260602.6	2.5	**48.0843**	0.971	2.0	ug/L	240	Standard
	Pb	207	216580.3	1.8	**44.9759**	0.652	1.4	ug/L	198	Standard
	Pb	208	918219.8	1.2	**48.2652**	0.290	0.6	ug/L	883	Standard
	U	238	196774.7	2.4	**28.2131**	0.460	1.6	ug/L	3	Standard
>	**Bi**	209	530811.0	1.1				ug/L	523984	Standard

Sample ID: L1606088803S WG573099-04
Report Date/Time: Wednesday, June 29, 2016 15:22:23
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**12.2205**	21.158	173.1	mg/L	0	Standard
Mg	24	365.0	3.6	**0.2314**	0.007	3.2	mg/L	37	Standard
K	39	1516.7	5.6	**1.9963**	0.137	6.9	mg/L	38	Standard
Ca	43	726.7	5.8	**26.3770**	1.992	7.6	mg/L	10	Standard
Fe	54	108682.6	3.2	**57.8341**	0.799	1.4	mg/L	179	Standard
Fe	57	27718.7	1.4	**58.2169**	0.892	1.5	mg/L	335	Standard
Sc-1	45	96992.9	2.3				mg/L	61029	Standard
Cl	35	26216.7	4.4				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	4344.0	8.1				ug/L	1547	Standard
P	31	22273.0	3.2				ug/L	15286	Standard
S	34	2523.5	5.1				ug/L	1572	Standard
Sr	88	263.3	11.4				ug/L	73	Standard
C	12	56.7	10.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	52.9				mg/L	3	Standard
Dy	164	111985.6	3.7				mg/L	6	Standard
Ho-1	165	68191.4	0.6				mg/L	10	Standard
Er	166	57569.3	4.1				mg/L	13	Standard
I	127	6513.1	8.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		118.857	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		107.993	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606088803S WG573099-04
Report Date/Time: Wednesday, June 29, 2016 15:22:23
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	103.507
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	101.303
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ni 60 Upper, S, EEE	Ni	60	

Approved: July 04, 2016

Cu 65 Upper, S, EEE	Cu	65
Zn 66 Upper, S, EEE	Zn	66
Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088803SD WG573099-05
Sample Date/Time: Wednesday, June 29, 2016 15:23:18
Number of Replicates: 3
Autosampler Position: 340
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86779.6	2.9				ug/L	74699	Standard
	Be	9	26783.7	4.5	**26.1889**	0.607	2.3	ug/L	8	Standard
	Al	27	17495490.1	2.1	**102.4852**	1.702	1.7	ug/L	1493	Standard
	Sc	45	84971.1	3.4				ug/L	61029	Standard
	Ti	47	11032.7	2.5	**25.6633**	0.769	3.0	ug/L	35	Standard
	V	51	492228.0	0.7	**57.5050**	0.764	1.3	ug/L	1435	Standard
	Cr	52	556079.4	1.4	**65.4064**	1.154	1.8	ug/L	7635	Standard
	Cr	53	67044.7	1.2	**63.8916**	0.827	1.3	ug/L	363	Standard
	Mn	55	16288704.5	1.7	**2071.2097**	47.470	2.3	ug/L	747	Standard
	Co	59	487642.9	2.1	**65.3439**	1.764	2.7	ug/L	234	Standard
	Ni	60	230286.3	2.0	**108.7653**	2.155	2.0	ug/L	215	Standard
	Cu	65	186555.0	1.4	**88.6352**	1.512	1.7	ug/L	104	Standard
	Zn	66	269449.8	1.5	**217.9309**	3.745	1.7	ug/L	363	Standard
>	**Ge**	72	559301.6	0.7				ug/L	495142	Standard
	As	75	32569.4	2.0	**25.8033**	0.648	2.5	ug/L	-50	Standard
	Se	82	2764.3	2.2	**22.1623**	0.630	2.8	ug/L	14	Standard
	Se-1	77	1950.8	4.1	**23.0256**	1.042	4.5	ug/L	42	Standard
>	**Ga**	71	55862.8	2.2				mg/L	42	Standard
	Rb	85	472783.4	2.1				ug/L	17	Standard
	Y	89	931018.8	1.5				ug/L	417098	Standard
>	**Rh**	103	36.7	47.9				ug/L	13	Standard
	Mo	98	833.5	3.5	**0.3003**	0.015	5.0	ug/L	11	Standard
	Ag	107	3972.2	0.3	**0.5833**	0.011	1.9	ug/L	68	Standard
	Cd	111	54996.2	0.6	**25.7204**	0.316	1.2	mg/L	2	Standard
	Cd	114	133681.6	2.4	**25.4888**	0.206	0.8	ug/L	12	Standard
>	**In**	115	600647.8	1.6				ug/L	587441	Standard
	Sn	118	466.7	16.3	**-0.0167**	0.012	70.0	ug/L	497	Standard
	Sb	123	3387.1	2.7	**0.7053**	0.008	1.2	ug/L	78	Standard
	Ba	135	467382.4	1.0	**223.3343**	2.175	1.0	ug/L	24	Standard
	Ce	140	1516729.6	1.2				ug/L	25	Standard
>	**Tb**	159	1007426.2	2.1				ug/L	939701	Standard
	Ho	165	39640.1	1.9				ug/L	10	Standard
	Tl	203	211811.1	1.3	**24.5598**	0.357	1.5	ug/L	4	Standard
	Tl	205	187359.9	1.8	**25.0339**	0.597	2.4	ug/L	2	Standard
	Pb	206	219015.7	1.1	**39.1225**	0.790	2.0	ug/L	240	Standard
	Pb	207	182856.8	0.5	**36.7608**	0.679	1.8	ug/L	198	Standard
	Pb	208	768036.7	1.8	**39.0813**	0.960	2.5	ug/L	883	Standard
	U	238	187429.9	1.2	**26.0258**	0.546	2.1	ug/L	3	Standard
>	**Bi**	209	548237.3	1.8				ug/L	523984	Standard

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 1

Approved: July 04, 2016

Na	23	6.7	43.3	**55.5842**	25.460	45.8	mg/L	0	Standard
Mg	24	348.3	15.5	**0.2571**	0.039	15.2	mg/L	37	Standard
K	39	1421.7	3.5	**2.1441**	0.098	4.5	mg/L	38	Standard
Ca	43	383.3	17.4	**14.9477**	3.560	23.8	mg/L	10	Standard
Fe	54	86673.1	2.4	**52.6645**	1.510	2.9	mg/L	179	Standard
Fe	57	21795.7	3.1	**52.2085**	3.369	6.5	mg/L	335	Standard
> **Sc-1**	45	84971.1	3.4				mg/L	61029	Standard
Cl	35	24892.4	3.3				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	3257.0	4.5				ug/L	1547	Standard
P	31	20836.0	2.5				ug/L	15286	Standard
S	34	2406.9	4.4				ug/L	1572	Standard
Sr	88	191.7	46.0				ug/L	73	Standard
C	12	76.7	15.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	20.0				mg/L	3	Standard
Dy	164	63526.1	1.7				mg/L	6	Standard
Ho-1	165	39640.1	1.9				mg/L	10	Standard
Er	166	35224.2	2.0				mg/L	13	Standard
I	127	5139.2	9.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		116.173	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		112.958	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	102.248
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.629
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ni 60 Upper, S, EEE	Ni	60	

Sample ID: L1606088803SD WG573099-05
Report Date/Time: Wednesday, June 29, 2016 15:25:34
Page 3

Approved: July 04, 2016

Zn 66 Upper, S, EEE	Zn	66
Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606088301
Sample Date/Time: Wednesday, June 29, 2016 15:26:29
Number of Replicates: 3
Autosampler Position: 341
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84384.4	1.7				ug/L	74699	Standard
	Be	9	1073.4	1.5	**1.0872**	0.032	3.0	ug/L	8	Standard
	Al	27	11940771.9	2.8	**71.9363**	2.792	3.9	ug/L	1493	Standard
	Sc	45	76626.6	1.2				ug/L	61029	Standard
	Ti	47	4374.3	2.0	**10.2834**	0.247	2.4	ug/L	35	Standard
	V	51	162703.8	1.0	**19.2014**	0.098	0.5	ug/L	1435	Standard
	Cr	52	132016.8	1.7	**15.0167**	0.347	2.3	ug/L	7635	Standard
	Cr	53	16153.6	0.8	**15.4073**	0.063	0.4	ug/L	363	Standard
	Mn	55	3262876.5	0.6	**421.8414**	5.005	1.2	ug/L	747	Standard
	Co	59	67832.8	0.7	**9.2120**	0.089	1.0	ug/L	234	Standard
	Ni	60	42623.2	1.3	**20.3743**	0.274	1.3	ug/L	215	Standard
	Cu	65	31205.8	0.3	**14.9935**	0.149	1.0	ug/L	104	Standard
	Zn	66	53134.7	1.9	**43.3429**	0.711	1.6	ug/L	363	Standard
>	**Ge**	72	549915.9	0.8				ug/L	495142	Standard
	As	75	3335.6	3.0	**2.7409**	0.080	2.9	ug/L	-50	Standard
	Se	82	46.6	19.0	**0.2449**	0.076	30.9	ug/L	14	Standard
	Se-1	77	111.7	5.1	**0.7542**	0.066	8.8	ug/L	42	Standard
>	**Ga**	71	17852.2	2.1				mg/L	42	Standard
	Rb	85	173572.1	2.0				ug/L	17	Standard
	Y	89	703667.4	2.2				ug/L	417098	Standard
>	**Rh**	103	33.3	34.6				ug/L	13	Standard
	Mo	98	357.4	3.8	**0.1158**	0.007	6.2	ug/L	11	Standard
	Ag	107	550.0	6.6	**0.0636**	0.007	11.1	ug/L	68	Standard
	Cd	111	273.3	3.2	**0.1105**	0.006	5.2	mg/L	2	Standard
	Cd	114	617.1	4.0	**0.1069**	0.006	5.7	ug/L	12	Standard
>	**In**	115	648944.8	3.3				ug/L	587441	Standard
	Sn	118	606.7	6.6	**-0.0008**	0.003	418.3	ug/L	497	Standard
	Sb	123	137.4	14.1	**0.0224**	0.003	13.8	ug/L	78	Standard
	Ba	135	318337.1	2.2	**140.8176**	2.386	1.7	ug/L	24	Standard
	Ce	140	1211648.5	1.9				ug/L	25	Standard
>	**Tb**	159	1041543.0	2.5				ug/L	939701	Standard
	Ho	165	19666.1	3.2				ug/L	10	Standard
	Tl	203	1035.0	3.0	**0.1143**	0.004	3.7	ug/L	4	Standard
	Tl	205	953.4	8.0	**0.1233**	0.009	7.5	ug/L	2	Standard
	Pb	206	127279.7	1.1	**21.9443**	0.442	2.0	ug/L	240	Standard
	Pb	207	103046.6	1.5	**19.9907**	0.426	2.1	ug/L	198	Standard
	Pb	208	441148.5	1.4	**21.6612**	0.196	0.9	ug/L	883	Standard
	U	238	5034.2	0.5	**0.6713**	0.005	0.8	ug/L	3	Standard
>	**Bi**	209	567168.2	1.1				ug/L	523984	Standard

Sample ID: L1606088301
Report Date/Time: Wednesday, June 29, 2016 15:28:46
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	340.0	3.9	**0.2844**	0.011	3.7	mg/L	37	Standard
K	39	688.3	2.3	**1.0975**	0.019	1.8	mg/L	38	Standard
Ca	43	145.0	15.8	**4.7718**	1.160	24.3	mg/L	10	Standard
Fe	54	17324.9	3.4	**11.5598**	0.337	2.9	mg/L	179	Standard
Fe	57	4550.7	3.6	**11.2444**	0.399	3.5	mg/L	335	Standard
Sc-1	45	76626.6	1.2				mg/L	61029	Standard
Cl	35	25558.2	1.3				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	2263.5	8.1				ug/L	1547	Standard
P	31	20323.6	0.9				ug/L	15286	Standard
S	34	2258.5	1.1				ug/L	1572	Standard
Sr	88	101.7	10.2				ug/L	73	Standard
C	12	116.7	13.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	49.5				mg/L	3	Standard
Dy	164	32996.6	2.0				mg/L	6	Standard
Ho-1	165	19666.1	3.2				mg/L	10	Standard
Er	166	17349.9	3.1				mg/L	13	Standard
I	127	11507.7	1.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		112.966	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		111.062	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606088301
Report Date/Time: Wednesday, June 29, 2016 15:28:46
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	110.470
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.241
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606088301
Report Date/Time: Wednesday, June 29, 2016 15:28:46
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201
Sample Date/Time: Wednesday, June 29, 2016 15:29:41
Number of Replicates: 3
Autosampler Position: 342
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84277.1	1.0				ug/L	74699	Standard
	Be	9	1593.4	6.0	**1.6114**	0.083	5.1	ug/L	8	Standard
	Al	27	15243679.1	1.9	**91.9299**	1.967	2.1	ug/L	1493	Standard
	Sc	45	83538.1	3.1				ug/L	61029	Standard
	Ti	47	10307.8	1.1	**23.4331**	0.360	1.5	ug/L	35	Standard
	V	51	410563.1	0.8	**46.8597**	0.438	0.9	ug/L	1435	Standard
	Cr	52	261972.4	0.3	**29.5747**	0.180	0.6	ug/L	7635	Standard
	Cr	53	32057.2	1.7	**29.6933**	0.417	1.4	ug/L	363	Standard
	Mn	55	12908980.1	3.2	**1604.9549**	58.354	3.6	ug/L	747	Standard
	Co	59	181241.6	3.1	**23.7226**	0.846	3.6	ug/L	234	Standard
	Ni	60	66685.8	0.9	**30.7063**	0.259	0.8	ug/L	215	Standard
	Cu	65	73374.2	0.7	**34.0220**	0.398	1.2	ug/L	104	Standard
	Zn	66	143517.2	0.9	**113.2755**	1.322	1.2	ug/L	363	Standard
>	**Ge**	72	572023.7	0.5				ug/L	495142	Standard
	As	75	16668.5	2.4	**12.9416**	0.353	2.7	ug/L	-50	Standard
	Se	82	187.7	6.4	**1.3421**	0.095	7.1	ug/L	14	Standard
	Se-1	77	232.0	7.5	**2.1278**	0.217	10.2	ug/L	42	Standard
>	**Ga**	71	37605.0	3.4				mg/L	42	Standard
	Rb	85	431124.7	1.0				ug/L	17	Standard
	Y	89	907846.1	0.7				ug/L	417098	Standard
>	**Rh**	103	28.3	27.0				ug/L	13	Standard
	Mo	98	1317.8	2.4	**0.4515**	0.009	1.9	ug/L	11	Standard
	Ag	107	1338.4	1.8	**0.1770**	0.004	2.5	ug/L	68	Standard
	Cd	111	803.3	6.8	**0.3473**	0.022	6.2	mg/L	2	Standard
	Cd	114	1888.8	7.2	**0.3384**	0.023	6.7	ug/L	12	Standard
>	**In**	115	635240.6	1.1				ug/L	587441	Standard
	Sn	118	608.3	2.5	**0.0016**	0.002	109.2	ug/L	497	Standard
	Sb	123	242.0	3.6	**0.0438**	0.002	4.9	ug/L	78	Standard
	Ba	135	645480.7	0.7	**291.6323**	1.817	0.6	ug/L	24	Standard
	Ce	140	2459220.7	0.8				ug/L	25	Standard
>	**Tb**	159	1043691.8	1.6				ug/L	939701	Standard
	Ho	165	34592.8	2.2				ug/L	10	Standard
	Tl	203	3164.7	1.5	**0.3540**	0.007	2.0	ug/L	4	Standard
	Tl	205	2843.6	5.4	**0.3684**	0.020	5.4	ug/L	2	Standard
	Pb	206	325793.2	0.7	**56.4391**	0.926	1.6	ug/L	240	Standard
	Pb	207	262124.0	0.3	**51.1005**	0.344	0.7	ug/L	198	Standard
	Pb	208	1122453.0	1.2	**55.3927**	1.161	2.1	ug/L	883	Standard
	U	238	18503.6	0.6	**2.4863**	0.037	1.5	ug/L	3	Standard
>	**Bi**	209	565563.6	0.9				ug/L	523984	Standard

Sample ID: L1606091201
Report Date/Time: Wednesday, June 29, 2016 15:31:58
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	148.3	19.2	**0.0751**	0.029	38.8	mg/L	37	Standard
K	39	1071.7	2.8	**1.6167**	0.038	2.4	mg/L	38	Standard
Ca	43	301.7	16.4	**11.4520**	2.661	23.2	mg/L	10	Standard
Fe	54	45977.7	1.1	**28.3556**	0.839	3.0	mg/L	179	Standard
Fe	57	11742.9	4.0	**28.0839**	0.746	2.7	mg/L	335	Standard
Sc-1	45	83538.1	3.1				mg/L	61029	Standard
Cl	35	25586.9	0.8				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	3940.5	1.8				ug/L	1547	Standard
P	31	20170.1	2.1				ug/L	15286	Standard
S	34	2405.2	3.2				ug/L	1572	Standard
Sr	88	111.7	6.8				ug/L	73	Standard
C	12	316.7	15.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	146.7	35.0				mg/L	3	Standard
Dy	164	57194.9	2.2				mg/L	6	Standard
Ho-1	165	34592.8	2.2				mg/L	10	Standard
Er	166	32892.3	1.1				mg/L	13	Standard
I	127	48256.8	4.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		112.823	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.527	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091201
Report Date/Time: Wednesday, June 29, 2016 15:31:58
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.137
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.935
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606091201
Report Date/Time: Wednesday, June 29, 2016 15:31:58
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201PS WG573717-03
Sample Date/Time: Wednesday, June 29, 2016 15:32:52
Number of Replicates: 3
Autosampler Position: 343
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	84925.7	1.3				ug/L	74699	Standard
	Be	9	54409.2	0.5	**54.3741**	0.919	1.7	ug/L	8	Standard
	Al	27	15619580.7	1.4	**93.4869**	2.147	2.3	ug/L	1493	Standard
	Sc	45	84914.0	2.3				ug/L	61029	Standard
	Ti	47	10489.6	2.4	**24.7035**	0.475	1.9	ug/L	35	Standard
	V	51	847646.1	0.7	**100.4344**	1.074	1.1	ug/L	1435	Standard
	Cr	52	696222.3	1.3	**83.2031**	0.580	0.7	ug/L	7635	Standard
	Cr	53	86425.9	2.6	**83.5032**	1.546	1.9	ug/L	363	Standard
	Mn	55	12964778.0	1.1	**1669.4256**	16.996	1.0	ug/L	747	Standard
	Co	59	568224.0	0.4	**77.1151**	0.744	1.0	ug/L	234	Standard
	Ni	60	176707.0	1.9	**84.4902**	0.918	1.1	ug/L	215	Standard
	Cu	65	181984.3	1.8	**87.5571**	0.864	1.0	ug/L	104	Standard
	Zn	66	208968.8	1.8	**171.0579**	1.957	1.1	ug/L	363	Standard
>	Ge	72	552256.5	0.8				ug/L	495142	Standard
	As	75	82371.4	1.6	**65.9929**	0.953	1.4	ug/L	-50	Standard
	Se	82	6597.3	1.6	**53.7542**	0.590	1.1	ug/L	14	Standard
	Se-1	77	4584.4	1.6	**55.6563**	1.034	1.9	ug/L	42	Standard
>	Ga	71	37593.3	1.5				mg/L	42	Standard
	Rb	85	428376.8	2.4				ug/L	17	Standard
	Y	89	890825.8	0.7				ug/L	417098	Standard
>	Rh	103	50.0	10.0				ug/L	13	Standard
	Mo	98	1335.6	2.9	**0.4503**	0.014	3.2	ug/L	11	Standard
	Ag	107	364007.1	1.3	**50.8126**	0.330	0.6	ug/L	68	Standard
	Cd	111	115695.5	1.2	**50.3468**	0.384	0.8	mg/L	2	Standard
	Cd	114	280904.1	1.9	**49.8365**	0.143	0.3	ug/L	12	Standard
>	In	115	645589.3	1.8				ug/L	587441	Standard
	Sn	118	755.0	4.3	**0.0229**	0.003	13.2	ug/L	497	Standard
	Sb	123	254065.1	1.3	**49.5173**	0.623	1.3	ug/L	78	Standard
	Ba	135	752068.0	1.2	**334.3566**	2.055	0.6	ug/L	24	Standard
	Ce	140	2468196.2	1.3				ug/L	25	Standard
>	Tb	159	1069312.5	1.3				ug/L	939701	Standard
	Ho	165	35102.3	1.1				ug/L	10	Standard
	Tl	203	458369.6	1.1	**50.4504**	0.349	0.7	ug/L	4	Standard
	Tl	205	409805.1	0.9	**51.9806**	1.259	2.4	ug/L	2	Standard
	Pb	206	619020.9	0.8	**105.0729**	1.076	1.0	ug/L	240	Standard
	Pb	207	511972.0	0.6	**97.8107**	1.381	1.4	ug/L	198	Standard
	Pb	208	2150459.1	0.4	**103.9876**	1.211	1.2	ug/L	883	Standard
	U	238	395808.0	0.4	**52.1735**	0.789	1.5	ug/L	3	Standard
>	Bi	209	577539.6	1.5				ug/L	523984	Standard

Sample ID: L1606091201PS WG573717-03
Report Date/Time: Wednesday, June 29, 2016 15:35:09
Page 1

Approved: July 04, 2016

Na	23	6.7	43.3	**54.9152**	22.487	40.9	mg/L	0	Standard
Mg	24	118.3	17.1	**0.0444**	0.016	36.3	mg/L	37	Standard
K	39	1026.7	7.6	**1.5167**	0.116	7.6	mg/L	38	Standard
Ca	43	296.7	14.3	**10.9789**	2.237	20.4	mg/L	10	Standard
Fe	54	45524.1	0.4	**27.6104**	0.565	2.0	mg/L	179	Standard
Fe	57	12291.7	1.9	**28.9717**	1.259	4.3	mg/L	335	Standard
> **Sc-1**	45	84914.0	2.3				mg/L	61029	Standard
Cl	35	25819.3	0.8				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	3963.9	6.7				ug/L	1547	Standard
P	31	20629.0	3.4				ug/L	15286	Standard
S	34	2280.2	5.7				ug/L	1572	Standard
Sr	88	103.3	12.2				ug/L	73	Standard
C	12	333.3	17.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	140.0	7.1				mg/L	3	Standard
Dy	164	56592.8	3.3				mg/L	6	Standard
Ho-1	165	35102.3	1.1				mg/L	10	Standard
Er	166	32397.9	1.1				mg/L	13	Standard
I	127	48815.3	4.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.691	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.535	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091201PS WG573717-03
Report Date/Time: Wednesday, June 29, 2016 15:35:09
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.899
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.221
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
V 51 Upper, S, EEE	V	51	
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606091201PS WG573717-03
Report Date/Time: Wednesday, June 29, 2016 15:35:09
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135
Pb 206 Upper, S, EEE	Pb	206
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201SDL WG573717-04
Sample Date/Time: Wednesday, June 29, 2016 15:36:04
Number of Replicates: 3
Autosampler Position: 344
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83450.9	2.1				ug/L	74699	Standard
	Be	9	378.3	16.7	**0.3916**	0.058	14.9	ug/L	8	Standard
	Al	27	3082760.3	2.5	**18.7625**	0.099	0.5	ug/L	1493	Standard
	Sc	45	72303.8	2.1				ug/L	61029	Standard
	Ti	47	2033.1	4.6	**4.7237**	0.243	5.1	ug/L	35	Standard
	V	51	79797.1	0.9	**9.3216**	0.052	0.6	ug/L	1435	Standard
	Cr	52	57744.9	2.6	**5.9945**	0.158	2.6	ug/L	7635	Standard
	Cr	53	6483.1	2.8	**5.9884**	0.155	2.6	ug/L	363	Standard
	Mn	55	2512242.8	1.3	**324.8710**	1.825	0.6	ug/L	747	Standard
	Co	59	34712.7	2.2	**4.6975**	0.071	1.5	ug/L	234	Standard
	Ni	60	13460.0	2.9	**6.3513**	0.158	2.5	ug/L	215	Standard
	Cu	65	15009.4	2.6	**7.1598**	0.141	2.0	ug/L	104	Standard
	Zn	66	29256.2	2.4	**23.6692**	0.389	1.6	ug/L	363	Standard
>	**Ge**	72	549686.2	0.8				ug/L	495142	Standard
	As	75	3156.6	2.0	**2.5979**	0.033	1.3	ug/L	-50	Standard
	Se	82	47.1	17.6	**0.2490**	0.070	28.2	ug/L	14	Standard
	Se-1	77	77.0	11.9	**0.3268**	0.106	32.4	ug/L	42	Standard
>	**Ga**	71	7136.7	1.1				mg/L	42	Standard
	Rb	85	84221.8	1.7				ug/L	17	Standard
	Y	89	569587.4	0.8				ug/L	417098	Standard
>	**Rh**	103	25.0	91.7				ug/L	13	Standard
	Mo	98	271.1	6.9	**0.0847**	0.006	6.6	ug/L	11	Standard
	Ag	107	378.0	10.8	**0.0387**	0.005	13.9	ug/L	68	Standard
	Cd	111	185.8	10.3	**0.0712**	0.009	12.3	mg/L	2	Standard
	Cd	114	406.5	6.5	**0.0684**	0.004	6.4	ug/L	12	Standard
>	**In**	115	660230.6	1.0				ug/L	587441	Standard
	Sn	118	501.7	8.5	**-0.0183**	0.006	32.3	ug/L	497	Standard
	Sb	123	1797.9	20.8	**0.3384**	0.070	20.8	ug/L	78	Standard
	Ba	135	129725.4	1.6	**56.3802**	0.972	1.7	ug/L	24	Standard
	Ce	140	480517.5	2.7				ug/L	25	Standard
>	**Tb**	159	1042248.3	0.8				ug/L	939701	Standard
	Ho	165	6823.2	4.6				ug/L	10	Standard
	Tl	203	727.7	7.7	**0.0794**	0.006	7.1	ug/L	4	Standard
	Tl	205	616.7	4.2	**0.0795**	0.002	2.4	ug/L	2	Standard
	Pb	206	64359.4	2.2	**11.0114**	0.128	1.2	ug/L	240	Standard
	Pb	207	51665.3	2.6	**9.9429**	0.134	1.3	ug/L	198	Standard
	Pb	208	222946.4	0.6	**10.8660**	0.247	2.3	ug/L	883	Standard
	U	238	3701.1	2.3	**0.4902**	0.006	1.1	ug/L	3	Standard
>	**Bi**	209	569638.6	2.0				ug/L	523984	Standard

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:38:21
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	61.7	28.5	**0.0019**	0.018	942.5	mg/L	37	Standard
K	39	203.3	18.5	**0.2649**	0.067	25.4	mg/L	38	Standard
Ca	43	75.0	26.7	**1.4632**	0.987	67.5	mg/L	10	Standard
Fe	54	9325.9	3.5	**6.5358**	0.204	3.1	mg/L	179	Standard
Fe	57	2630.2	2.1	**6.4718**	0.055	0.8	mg/L	335	Standard
> **Sc-1**	45	72303.8	2.1				mg/L	61029	Standard
Cl	35	24972.6	0.9				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2286.8	8.4				ug/L	1547	Standard
P	31	19043.6	4.4				ug/L	15286	Standard
S	34	2180.2	1.2				ug/L	1572	Standard
Sr	88	81.7	35.9				ug/L	73	Standard
C	12	93.3	22.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	57.3				mg/L	3	Standard
Dy	164	11677.0	2.4				mg/L	6	Standard
Ho-1	165	6823.2	4.6				mg/L	10	Standard
Er	166	6424.7	2.9				mg/L	13	Standard
I	127	12089.8	7.2				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		111.717	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		111.016	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:38:21
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.391
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.713
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:38:21
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606091201SDL WG573717-04
Sample Date/Time: Wednesday, June 29, 2016 15:39:15
Number of Replicates: 3
Autosampler Position: 345
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80082.8	1.5				ug/L	74699	Standard
	Be	9	80.0	10.8	**0.0924**	0.009	10.2	ug/L	8	Standard
	Al	27	638793.5	0.9	**4.0416**	0.057	1.4	ug/L	1493	Standard
	Sc	45	66977.8	2.5				ug/L	61029	Standard
	Ti	47	422.3	5.3	**0.8899**	0.044	4.9	ug/L	35	Standard
	V	51	17807.2	0.8	**1.9190**	0.038	2.0	ug/L	1435	Standard
	Cr	52	18051.7	1.5	**1.1605**	0.015	1.3	ug/L	7635	Standard
	Cr	53	1846.8	4.0	**1.4619**	0.067	4.6	ug/L	363	Standard
	Mn	55	540671.4	1.9	**69.4959**	1.549	2.2	ug/L	747	Standard
	Co	59	7520.2	3.3	**0.9841**	0.038	3.8	ug/L	234	Standard
	Ni	60	2912.3	1.6	**1.2710**	0.019	1.5	ug/L	215	Standard
	Cu	65	3203.0	1.2	**1.4395**	0.003	0.2	ug/L	104	Standard
	Zn	66	8550.1	1.3	**6.5648**	0.083	1.3	ug/L	363	Standard
>	**Ge**	72	552153.4	1.0				ug/L	495142	Standard
	As	75	624.0	5.3	**0.5591**	0.024	4.4	ug/L	-50	Standard
	Se	82	21.8	25.1	**0.0396**	0.043	108.7	ug/L	14	Standard
	Se-1	77	57.0	8.0	**0.0772**	0.050	64.2	ug/L	42	Standard
>	**Ga**	71	1470.1	2.9				mg/L	42	Standard
	Rb	85	17425.0	3.8				ug/L	17	Standard
	Y	89	497203.9	1.3				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	65.7	10.1	**0.0165**	0.003	15.8	ug/L	11	Standard
	Ag	107	156.7	8.7	**0.0087**	0.002	18.0	ug/L	68	Standard
	Cd	111	38.9	11.8	**0.0089**	0.002	23.3	mg/L	2	Standard
	Cd	114	108.8	14.4	**0.0170**	0.003	15.1	ug/L	12	Standard
>	**In**	115	651862.8	1.7				ug/L	587441	Standard
	Sn	118	480.0	4.2	**-0.0206**	0.003	16.7	ug/L	497	Standard
	Sb	123	490.3	31.4	**0.0908**	0.031	34.3	ug/L	78	Standard
	Ba	135	26415.7	0.1	**11.6170**	0.204	1.8	ug/L	24	Standard
	Ce	140	98329.7	0.4				ug/L	25	Standard
>	**Tb**	159	1007324.2	1.5				ug/L	939701	Standard
	Ho	165	1405.1	5.7				ug/L	10	Standard
	Tl	203	169.0	1.6	**0.0177**	0.000	0.8	ug/L	4	Standard
	Tl	205	175.0	8.6	**0.0234**	0.002	8.3	ug/L	2	Standard
	Pb	206	13505.0	1.4	**2.3274**	0.057	2.5	ug/L	240	Standard
	Pb	207	10900.6	1.3	**2.1089**	0.044	2.1	ug/L	198	Standard
	Pb	208	47029.6	1.3	**2.3051**	0.014	0.6	ug/L	883	Standard
	U	238	743.4	3.4	**0.0980**	0.002	2.1	ug/L	3	Standard
>	**Bi**	209	552389.3	1.4				ug/L	523984	Standard

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:41:32
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	33.3	17.3	**-0.0260**	0.006	22.5	mg/L	37	Standard
K	39	50.0	36.1	**-0.0133**	0.037	281.6	mg/L	38	Standard
Ca	43	41.7	48.5	**-0.1354**	1.146	846.9	mg/L	10	Standard
Fe	54	2020.3	1.1	**1.4233**	0.051	3.6	mg/L	179	Standard
Fe	57	728.4	10.0	**1.1787**	0.168	14.2	mg/L	335	Standard
> **Sc-1**	45	66977.8	2.5				mg/L	61029	Standard
Cl	35	24720.8	1.1				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	1860.1	4.8				ug/L	1547	Standard
P	31	18349.4	1.0				ug/L	15286	Standard
S	34	1845.1	4.4				ug/L	1572	Standard
Sr	88	65.0	27.7				ug/L	73	Standard
C	12	66.7	31.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	2304.6	5.8				mg/L	6	Standard
Ho-1	165	1405.1	5.7				mg/L	10	Standard
Er	166	1370.1	4.1				mg/L	13	Standard
I	127	4489.0	15.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		107.208	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.514	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606091201SDL WG573717-04
Report Date/Time: Wednesday, June 29, 2016 15:41:32
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.966
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.421
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 15:42:28
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79858.3	1.9				ug/L	74699	Standard
	Be	9	50198.1	2.5	**53.3484**	1.401	2.6	ug/L	8	Standard
	Al	27	8066966.3	0.3	**51.3427**	0.865	1.7	ug/L	1493	Standard
	Sc	45	68357.2	2.3				ug/L	61029	Standard
	Ti	47	43587.6	0.3	**105.7066**	1.724	1.6	ug/L	35	Standard
	V	51	433103.8	0.8	**52.5653**	0.576	1.1	ug/L	1435	Standard
	Cr	52	431112.3	0.9	**52.4990**	0.789	1.5	ug/L	7635	Standard
	Cr	53	52653.0	1.0	**52.0862**	0.689	1.3	ug/L	363	Standard
	Mn	55	405815.5	1.7	**53.4827**	1.214	2.3	ug/L	747	Standard
	Co	59	390955.7	1.0	**54.4418**	1.354	2.5	ug/L	234	Standard
	Ni	60	105213.1	1.7	**51.5742**	0.488	0.9	ug/L	215	Standard
	Cu	65	104377.6	0.2	**51.4980**	0.905	1.8	ug/L	104	Standard
	Zn	66	61654.9	0.7	**51.4853**	1.289	2.5	ug/L	363	Standard
>	**Ge**	72	538231.7	1.8				ug/L	495142	Standard
	As	75	61802.6	0.2	**50.8282**	0.881	1.7	ug/L	-50	Standard
	Se	82	6211.6	1.2	**51.9313**	0.364	0.7	ug/L	14	Standard
	Se-1	77	4323.6	2.0	**53.8508**	1.659	3.1	ug/L	42	Standard
>	**Ga**	71	60.0	28.9				mg/L	42	Standard
	Rb	85	955.0	5.2				ug/L	17	Standard
	Y	89	475770.1	1.3				ug/L	417098	Standard
>	**Rh**	103	25.0	34.6				ug/L	13	Standard
	Mo	98	290643.5	0.8	**97.6712**	0.653	0.7	ug/L	11	Standard
	Ag	107	354998.2	1.8	**48.7852**	1.288	2.6	ug/L	68	Standard
	Cd	111	115105.3	1.2	**49.3038**	0.151	0.3	mg/L	2	Standard
	Cd	114	278592.4	1.8	**48.6531**	0.360	0.7	ug/L	12	Standard
>	**In**	115	655825.0	1.0				ug/L	587441	Standard
	Sn	118	319548.2	1.2	**48.9035**	0.256	0.5	ug/L	497	Standard
	Sb	123	249456.0	1.2	**47.8604**	0.919	1.9	ug/L	78	Standard
	Ba	135	108215.4	0.8	**47.3455**	0.546	1.2	ug/L	24	Standard
	Ce	140	256.7	76.3				ug/L	25	Standard
>	**Tb**	159	1025574.2	1.9				ug/L	939701	Standard
	Ho	165	15.0	0.0				ug/L	10	Standard
	Tl	203	443897.9	1.2	**50.1327**	1.058	2.1	ug/L	4	Standard
	Tl	205	398087.0	0.7	**51.8003**	0.814	1.6	ug/L	2	Standard
	Pb	206	280554.3	0.5	**48.8226**	0.728	1.5	ug/L	240	Standard
	Pb	207	249585.9	0.6	**48.8846**	0.533	1.1	ug/L	198	Standard
	Pb	208	1026713.8	0.6	**50.9050**	1.131	2.2	ug/L	883	Standard
	U	238	372550.1	0.8	**50.3866**	1.127	2.2	ug/L	3	Standard
>	**Bi**	209	562933.7	1.7				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:44:45
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4292.3	1.6	**4.8821**	0.035	0.7	mg/L	37	Standard
K	39	2345.2	3.1	**4.5133**	0.080	1.8	mg/L	38	Standard
Ca	43	103.3	10.1	**3.3018**	0.528	16.0	mg/L	10	Standard
Fe	54	6681.2	2.2	**4.9206**	0.153	3.1	mg/L	179	Standard
Fe	57	1928.5	4.0	**4.7824**	0.344	7.2	mg/L	335	Standard
> **Sc-1**	45	68357.2	2.3				mg/L	61029	Standard
Cl	35	26753.6	0.7				ug/L	20195	Standard
Kr	83	3.3	45.8				ug/L	4	Standard
Br	81	1833.4	8.1				ug/L	1547	Standard
P	31	20395.4	2.2				ug/L	15286	Standard
S	34	2408.5	8.3				ug/L	1572	Standard
Sr	88	85.0	25.6				ug/L	73	Standard
C	12	53.3	28.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.2	35.5				mg/L	6	Standard
Ho-1	165	15.0	0.0				mg/L	10	Standard
Er	166	30.0	57.7				mg/L	13	Standard
I	127	2536.9	16.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	106.697		
Al	27	102.685		
Sc	45			
Ti	47	105.707		
V	51	105.131		
Cr	52	104.998		
Cr	53			
Mn	55	106.965		
Co	59	108.884		
Ni	60	103.148		
Cu	65	102.996		
Zn	66	102.971		
> Ge	72		108.702	
As	75	101.656		
Se	82	103.863		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:44:45
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	97.671		
Ag	107	97.570		
Cd	111	98.608		
Cd	114			
In	115			111.641
Sn	118	97.807		
Sb	123	95.721		
Ba	135	94.691		
Ce	140			
Tb	159			
Ho	165			
Tl	203	100.265		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.810		
U	238	100.773		
Bi	209			107.433
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 15:44:45
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 15:45:40
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	76618.3	2.3				ug/L	74699	Standard
	Be	9	25.0	40.0	**0.0352**	0.010	29.8	ug/L	8	Standard
	Al	27	2451.9	23.9	**0.0029**	0.004	131.5	ug/L	1493	Standard
	Sc	45	64234.2	1.5				ug/L	61029	Standard
	Ti	47	38.0	9.5	**-0.0165**	0.010	60.4	ug/L	35	Standard
	V	51	1363.5	4.5	**-0.0292**	0.006	19.9	ug/L	1435	Standard
	Cr	52	7102.3	0.4	**-0.1401**	0.016	11.4	ug/L	7635	Standard
	Cr	53	373.3	16.1	**0.0429**	0.063	148.1	ug/L	363	Standard
	Mn	55	1097.0	20.0	**-0.0025**	0.031	1279.7	ug/L	747	Standard
	Co	59	247.3	6.4	**-0.0027**	0.003	101.3	ug/L	234	Standard
	Ni	60	180.3	4.0	**-0.0351**	0.003	7.6	ug/L	215	Standard
	Cu	65	132.0	9.2	**-0.0382**	0.007	18.6	ug/L	104	Standard
	Zn	66	415.3	4.4	**-0.1038**	0.019	18.2	ug/L	363	Standard
>	Ge	72	537099.7	1.7				ug/L	495142	Standard
	As	75	-44.5	23.9	**0.0230**	0.008	35.8	ug/L	-50	Standard
	Se	82	14.3	4.3	**-0.0180**	0.004	23.6	ug/L	14	Standard
	Se-1	77	59.7	7.9	**0.1303**	0.047	35.8	ug/L	42	Standard
>	Ga	71	51.7	29.6				mg/L	42	Standard
	Rb	85	35.0	0.0				ug/L	17	Standard
	Y	89	479758.9	0.4				ug/L	417098	Standard
>	Rh	103	8.3	34.6				ug/L	13	Standard
	Mo	98	277.2	29.2	**0.0885**	0.026	29.3	ug/L	11	Standard
	Ag	107	131.7	17.8	**0.0055**	0.003	58.6	ug/L	68	Standard
	Cd	111	17.2	29.5	**-0.0005**	0.002	465.0	mg/L	2	Standard
	Cd	114	40.7	69.7	**0.0051**	0.005	98.6	ug/L	12	Standard
>	In	115	646048.6	1.6				ug/L	587441	Standard
	Sn	118	843.4	4.5	**0.0366**	0.007	19.0	ug/L	497	Standard
	Sb	123	501.9	29.2	**0.0939**	0.030	32.1	ug/L	78	Standard
	Ba	135	52.0	15.0	**0.0071**	0.003	45.6	ug/L	24	Standard
	Ce	140	70.0	58.5				ug/L	25	Standard
>	Tb	159	979678.9	1.9				ug/L	939701	Standard
	Ho	165	13.3	114.6				ug/L	10	Standard
	Tl	203	68.0	38.6	**0.0062**	0.003	47.0	ug/L	4	Standard
	Tl	205	56.7	27.0	**0.0078**	0.002	25.4	ug/L	2	Standard
	Pb	206	311.7	5.0	**-0.0149**	0.002	16.8	ug/L	240	Standard
	Pb	207	257.0	6.3	**-0.0177**	0.002	12.3	ug/L	198	Standard
	Pb	208	1140.4	10.1	**-0.0154**	0.005	31.9	ug/L	883	Standard
	U	238	55.7	48.1	**0.0033**	0.004	107.8	ug/L	3	Standard
>	Bi	209	542619.7	2.4				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:47:57
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.9504**	31.082	173.2	mg/L	0	Standard
Mg	24	40.0	54.5	**-0.0162**	0.026	161.8	mg/L	37	Standard
K	39	53.3	59.5	**-0.0020**	0.067	3445.3	mg/L	38	Standard
Ca	43	18.3	103.3	**-1.4157**	1.156	81.6	mg/L	10	Standard
Fe	54	130.7	11.8	**-0.0329**	0.014	41.5	mg/L	179	Standard
Fe	57	301.7	9.1	**-0.1005**	0.081	80.3	mg/L	335	Standard
> **Sc-1**	45	64234.2	1.5				mg/L	61029	Standard
Cl	35	26295.5	2.8				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1633.4	7.7				ug/L	1547	Standard
P	31	19669.4	2.1				ug/L	15286	Standard
S	34	2230.2	5.1				ug/L	1572	Standard
Sr	88	70.0	21.4				ug/L	73	Standard
C	12	46.7	61.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	12.2	49.4				mg/L	6	Standard
Ho-1	165	13.3	114.6				mg/L	10	Standard
Er	166	23.3	107.9				mg/L	13	Standard
I	127	1980.1	22.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.474	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:47:57
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.977
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.556
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 15:47:57
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBW 60 WG574001-04
Sample Date/Time: Wednesday, June 29, 2016 15:48:53
Number of Replicates: 3
Autosampler Position: 205
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79282.0	2.9				ug/L	74699	Standard
	Be	9	31.7	134.3	**0.0407**	0.044	108.3	ug/L	8	Standard
	Al	27	19345.2	59.5	**0.1093**	0.069	63.4	ug/L	1493	Standard
	Sc	45	67500.3	4.8				ug/L	61029	Standard
	Ti	47	51.7	27.5	**0.0181**	0.032	177.8	ug/L	35	Standard
	V	51	1877.4	33.1	**0.0362**	0.072	198.6	ug/L	1435	Standard
	Cr	52	9295.2	4.7	**0.1500**	0.032	21.3	ug/L	7635	Standard
	Cr	53	908.4	6.4	**0.5905**	0.063	10.7	ug/L	363	Standard
	Mn	55	10971.3	111.8	**1.3060**	1.603	122.8	ug/L	747	Standard
	Co	59	555.3	78.4	**0.0407**	0.060	146.5	ug/L	234	Standard
	Ni	60	485.3	47.2	**0.1173**	0.109	92.6	ug/L	215	Standard
	Cu	65	307.3	25.1	**0.0504**	0.035	70.1	ug/L	104	Standard
	Zn	66	1901.5	5.5	**1.1744**	0.056	4.8	ug/L	363	Standard
>	**Ge**	72	528898.5	2.0				ug/L	495142	Standard
	As	75	-0.1	63053.1	**0.0590**	0.043	73.0	ug/L	-50	Standard
	Se	82	20.7	41.6	**0.0373**	0.070	187.2	ug/L	14	Standard
	Se-1	77	69.7	5.4	**0.2713**	0.061	22.4	ug/L	42	Standard
>	**Ga**	71	70.0	75.3				mg/L	42	Standard
	Rb	85	391.7	126.9				ug/L	17	Standard
	Y	89	488721.1	2.3				ug/L	417098	Standard
>	**Rh**	103	16.7	105.4				ug/L	13	Standard
	Mo	98	140.7	71.4	**0.0402**	0.032	80.6	ug/L	11	Standard
	Ag	107	241.3	88.8	**0.0193**	0.028	147.0	ug/L	68	Standard
	Cd	111	97.4	139.8	**0.0324**	0.056	173.4	mg/L	2	Standard
	Cd	114	250.4	134.1	**0.0402**	0.057	140.7	ug/L	12	Standard
>	**In**	115	673056.2	1.0				ug/L	587441	Standard
	Sn	118	918.4	10.7	**0.0424**	0.014	32.3	ug/L	497	Standard
	Sb	123	377.5	33.3	**0.0663**	0.023	34.5	ug/L	78	Standard
	Ba	135	710.7	130.6	**0.2849**	0.392	137.5	ug/L	24	Standard
	Ce	140	2190.4	144.6				ug/L	25	Standard
>	**Tb**	159	1033072.6	2.9				ug/L	939701	Standard
	Ho	165	46.7	118.0				ug/L	10	Standard
	Tl	203	354.0	129.7	**0.0375**	0.050	134.0	ug/L	4	Standard
	Tl	205	175.0	119.3	**0.0226**	0.026	116.3	ug/L	2	Standard
	Pb	206	584.7	65.9	**0.0302**	0.064	211.0	ug/L	240	Standard
	Pb	207	508.0	67.6	**0.0291**	0.064	219.9	ug/L	198	Standard
	Pb	208	1976.7	53.7	**0.0238**	0.049	207.9	ug/L	883	Standard
	U	238	147.7	117.1	**0.0152**	0.023	148.4	ug/L	3	Standard
>	**Bi**	209	558931.4	2.9				ug/L	523984	Standard

Sample ID: PBW 60 WG574001-04
Report Date/Time: Wednesday, June 29, 2016 15:51:10
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	38.3	32.8	**-0.0206**	0.014	67.5	mg/L	37	Standard
K	39	46.7	37.6	**-0.0214**	0.032	149.6	mg/L	38	Standard
Ca	43	20.0	0.0	**-1.3794**	0.055	4.0	mg/L	10	Standard
Fe	54	191.0	21.0	**0.0073**	0.024	325.9	mg/L	179	Standard
Fe	57	261.7	19.2	**-0.2718**	0.132	48.6	mg/L	335	Standard
> **Sc-1**	45	67500.3	4.8				mg/L	61029	Standard
Cl	35	25962.2	2.0				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	2216.8	6.9				ug/L	1547	Standard
P	31	19524.2	2.3				ug/L	15286	Standard
S	34	2248.5	4.1				ug/L	1572	Standard
Sr	88	73.3	27.6				ug/L	73	Standard
C	12	46.7	44.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	67.1	124.3				mg/L	6	Standard
Ho-1	165	46.7	118.0				mg/L	10	Standard
Er	166	60.0	130.2				mg/L	13	Standard
I	127	1978.5	35.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		106.136	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		106.817	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBW 60 WG574001-04
Report Date/Time: Wednesday, June 29, 2016 15:51:10
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.574
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.669
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 60 WG574001-04
Report Date/Time: Wednesday, June 29, 2016 15:51:10
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSW 60 WG574001-05
Sample Date/Time: Wednesday, June 29, 2016 15:52:04
Number of Replicates: 3
Autosampler Position: 206
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78008.7	2.7				ug/L	74699	Standard
	Be	9	49798.4	3.1	**54.1693**	0.583	1.1	ug/L	8	Standard
	Al	27	8265.6	3.2	**0.0405**	0.001	2.7	ug/L	1493	Standard
	Sc	45	64878.8	4.3				ug/L	61029	Standard
	Ti	47	43.0	16.3	**-0.0042**	0.019	444.7	ug/L	35	Standard
	V	51	424123.5	0.9	**51.5862**	0.653	1.3	ug/L	1435	Standard
	Cr	52	430096.1	0.4	**52.4970**	1.185	2.3	ug/L	7635	Standard
	Cr	53	53307.0	1.3	**52.8599**	1.230	2.3	ug/L	363	Standard
	Mn	55	405428.0	0.3	**53.5518**	1.000	1.9	ug/L	747	Standard
	Co	59	387903.4	0.1	**54.1349**	1.154	2.1	ug/L	234	Standard
	Ni	60	106395.6	1.4	**52.2826**	1.359	2.6	ug/L	215	Standard
	Cu	65	107418.5	1.1	**53.1215**	1.283	2.4	ug/L	104	Standard
	Zn	66	63602.2	0.4	**53.2433**	1.221	2.3	ug/L	363	Standard
>	**Ge**	72	537059.3	2.0				ug/L	495142	Standard
	As	75	62286.8	1.3	**51.3419**	1.295	2.5	ug/L	-50	Standard
	Se	82	6132.0	1.6	**51.3807**	0.931	1.8	ug/L	14	Standard
	Se-1	77	4215.6	1.2	**52.6015**	0.951	1.8	ug/L	42	Standard
>	**Ga**	71	76.7	32.8				mg/L	42	Standard
	Rb	85	50.0	36.1				ug/L	17	Standard
	Y	89	473307.8	1.1				ug/L	417098	Standard
>	**Rh**	103	23.3	75.3				ug/L	13	Standard
	Mo	98	59.9	20.8	**0.0148**	0.004	26.5	ug/L	11	Standard
	Ag	107	358191.0	0.9	**50.3679**	1.706	3.4	ug/L	68	Standard
	Cd	111	112882.0	2.3	**49.4684**	1.332	2.7	mg/L	2	Standard
	Cd	114	273909.8	1.0	**48.9526**	1.484	3.0	ug/L	12	Standard
>	**In**	115	641215.1	2.7				ug/L	587441	Standard
	Sn	118	785.0	7.1	**0.0286**	0.011	37.9	ug/L	497	Standard
	Sb	123	249794.4	1.0	**49.0383**	1.564	3.2	ug/L	78	Standard
	Ba	135	107708.9	1.4	**48.2176**	1.451	3.0	ug/L	24	Standard
	Ce	140	101.7	27.1				ug/L	25	Standard
>	**Tb**	159	990038.9	2.7				ug/L	939701	Standard
	Ho	165	11.7	99.0				ug/L	10	Standard
	Tl	203	439976.8	0.4	**51.4932**	0.969	1.9	ug/L	4	Standard
	Tl	205	398605.3	0.6	**53.7491**	0.791	1.5	ug/L	2	Standard
	Pb	206	290681.6	0.4	**52.4241**	0.636	1.2	ug/L	240	Standard
	Pb	207	247225.7	0.4	**50.1841**	0.852	1.7	ug/L	198	Standard
	Pb	208	1030904.4	0.6	**52.9701**	1.178	2.2	ug/L	883	Standard
	U	238	368133.3	0.5	**51.5924**	0.851	1.6	ug/L	3	Standard
>	**Bi**	209	543222.2	1.6				ug/L	523984	Standard

Sample ID: LCSW 60 WG574001-05
Report Date/Time: Wednesday, June 29, 2016 15:54:21
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	35.0	28.6	**-0.0227**	0.012	52.0	mg/L	37	Standard
K	39	28.3	50.9	**-0.0559**	0.027	48.5	mg/L	38	Standard
Ca	43	35.0	28.6	**-0.4579**	0.506	110.5	mg/L	10	Standard
Fe	54	138.7	22.6	**-0.0277**	0.025	88.8	mg/L	179	Standard
Fe	57	243.3	16.6	**-0.2916**	0.161	55.3	mg/L	335	Standard
> **Sc-1**	45	64878.8	4.3				mg/L	61029	Standard
Cl	35	24949.9	2.3				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	1880.1	6.1				ug/L	1547	Standard
P	31	18362.8	1.6				ug/L	15286	Standard
S	34	2046.8	4.1				ug/L	1572	Standard
Sr	88	70.0	18.9				ug/L	73	Standard
C	12	56.7	56.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	2.2	249.0				mg/L	6	Standard
Ho-1	165	11.7	99.0				mg/L	10	Standard
Er	166	23.3	49.5				mg/L	13	Standard
I	127	1781.8	25.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		104.431	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.466	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 60 WG574001-05
Report Date/Time: Wednesday, June 29, 2016 15:54:21
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.154
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.671
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 60 WG574001-05
Report Date/Time: Wednesday, June 29, 2016 15:54:21
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124702 WG574001-01
Sample Date/Time: Wednesday, June 29, 2016 15:55:16
Number of Replicates: 3
Autosampler Position: 207
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	134635.3	6.0				ug/L	74699	Standard
	Be	9	80.0	18.8	**0.0585**	0.013	21.4	ug/L	8	Standard
	Al	27	38117257.0	1.5	**144.1796**	7.359	5.1	ug/L	1493	Standard
	Sc	45	57575.9	1.8				ug/L	61029	Standard
	Ti	47	164.0	8.6	**0.3117**	0.039	12.5	ug/L	35	Standard
	V	51	2694.7	7.4	**0.1512**	0.027	17.7	ug/L	1435	Standard
	Cr	52	13340.2	1.7	**0.7256**	0.017	2.4	ug/L	7635	Standard
	Cr	53	1556.7	8.2	**1.3078**	0.145	11.1	ug/L	363	Standard
	Mn	55	855520.0	4.5	**119.2592**	4.647	3.9	ug/L	747	Standard
	Co	59	4433.3	4.9	**0.6151**	0.027	4.4	ug/L	234	Standard
	Ni	60	11360.3	1.4	**5.7737**	0.123	2.1	ug/L	215	Standard
	Cu	65	4582.0	1.3	**2.2889**	0.015	0.7	ug/L	104	Standard
	Zn	66	3866.8	1.4	**2.9861**	0.052	1.7	ug/L	363	Standard
>	**Ge**	72	509457.0	0.7				ug/L	495142	Standard
	As	75	332.4	10.8	**0.3479**	0.030	8.6	ug/L	-50	Standard
	Se	82	161.9	7.9	**1.2952**	0.108	8.4	ug/L	14	Standard
	Se-1	77	102.3	14.4	**0.7391**	0.192	26.0	ug/L	42	Standard
>	**Ga**	71	95.0	18.2				mg/L	42	Standard
	Rb	85	141592.2	2.6				ug/L	17	Standard
	Y	89	417616.8	3.6				ug/L	417098	Standard
>	**Rh**	103	198.3	13.9				ug/L	13	Standard
	Mo	98	173.0	3.6	**0.0623**	0.003	5.6	ug/L	11	Standard
	Ag	107	339.3	7.8	**0.0417**	0.005	13.0	ug/L	68	Standard
	Cd	111	139.7	8.2	**0.0622**	0.007	10.5	mg/L	2	Standard
	Cd	114	358.4	9.8	**0.0712**	0.009	12.2	ug/L	12	Standard
>	**In**	115	560534.9	2.1				ug/L	587441	Standard
	Sn	118	620.0	5.6	**0.0166**	0.008	49.6	ug/L	497	Standard
	Sb	123	641.7	15.5	**0.1401**	0.024	17.2	ug/L	78	Standard
	Ba	135	5098.9	2.7	**2.5958**	0.100	3.9	ug/L	24	Standard
	Ce	140	815.0	33.2				ug/L	25	Standard
>	**Tb**	159	928792.3	1.3				ug/L	939701	Standard
	Ho	165	11.7	65.5				ug/L	10	Standard
	Tl	203	1143.7	8.4	**0.1440**	0.014	9.4	ug/L	4	Standard
	Tl	205	1088.4	7.4	**0.1600**	0.012	7.4	ug/L	2	Standard
	Pb	206	578.7	9.2	**0.0425**	0.010	23.1	ug/L	240	Standard
	Pb	207	493.0	9.5	**0.0393**	0.012	31.8	ug/L	198	Standard
	Pb	208	1938.4	3.0	**0.0345**	0.004	12.8	ug/L	883	Standard
	U	238	1219.0	4.3	**0.1816**	0.009	5.1	ug/L	3	Standard
>	**Bi**	209	499083.1	3.3				ug/L	523984	Standard

Sample ID: L1606124702 WG574001-01
Report Date/Time: Wednesday, June 29, 2016 15:57:33
Page 1

Approved: July 04, 2016

Na	23	20.0	43.3	**245.3647**	110.079	44.9	mg/L	0	Standard
Mg	24	80086.1	0.7	**109.5385**	2.239	2.0	mg/L	37	Standard
K	39	1376.7	6.5	**3.1098**	0.163	5.2	mg/L	38	Standard
Ca	43	2828.6	6.5	**186.7611**	8.963	4.8	mg/L	10	Standard
Fe	54	188.7	25.7	**0.0318**	0.047	146.5	mg/L	179	Standard
Fe	57	1751.8	4.9	**5.2348**	0.203	3.9	mg/L	335	Standard
> **Sc-1**	45	57575.9	1.8				mg/L	61029	Standard
Cl	35	65110.3	1.2				ug/L	20195	Standard
Kr	83	3.7	68.6				ug/L	4	Standard
Br	81	22346.5	6.5				ug/L	1547	Standard
P	31	17116.3	3.4				ug/L	15286	Standard
S	34	2223.5	6.9				ug/L	1572	Standard
Sr	88	836.7	13.9				ug/L	73	Standard
C	12	76.7	74.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	52.1	77.4				mg/L	6	Standard
Ho-1	165	11.7	65.5				mg/L	10	Standard
Er	166	26.7	57.3				mg/L	13	Standard
I	127	41219.6	7.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		180.238	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		102.891	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606124702 WG574001-01
Report Date/Time: Wednesday, June 29, 2016 15:57:33
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.420
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		95.248
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124702 WG574001-01
Report Date/Time: Wednesday, June 29, 2016 15:57:33
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124702DP WG574001-06
Sample Date/Time: Wednesday, June 29, 2016 15:58:27
Number of Replicates: 3
Autosampler Position: 208
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	151702.5	2.0				ug/L	74699	Standard
	Be	9	51.7	59.1	**0.0368**	0.018	48.5	ug/L	8	Standard
	Al	27	41839885.7	0.1	**140.2128**	2.854	2.0	ug/L	1493	Standard
	Sc	45	56811.4	3.1				ug/L	61029	Standard
	Ti	47	131.7	6.1	**0.2545**	0.026	10.2	ug/L	35	Standard
	V	51	2376.3	9.9	**0.1332**	0.026	19.9	ug/L	1435	Standard
	Cr	52	12951.9	1.5	**0.8048**	0.052	6.4	ug/L	7635	Standard
	Cr	53	1565.1	1.7	**1.4417**	0.042	2.9	ug/L	363	Standard
	Mn	55	867683.1	1.0	**130.2493**	3.502	2.7	ug/L	747	Standard
	Co	59	4344.6	3.7	**0.6509**	0.022	3.4	ug/L	234	Standard
	Ni	60	12342.7	0.7	**6.7729**	0.126	1.9	ug/L	215	Standard
	Cu	65	4552.0	1.4	**2.4548**	0.036	1.5	ug/L	104	Standard
	Zn	66	4125.6	2.3	**3.4960**	0.055	1.6	ug/L	363	Standard
>	**Ge**	72	473381.7	2.0				ug/L	495142	Standard
	As	75	389.2	1.9	**0.4231**	0.002	0.6	ug/L	-50	Standard
	Se	82	181.1	4.2	**1.5888**	0.099	6.2	ug/L	14	Standard
	Se-1	77	106.3	18.0	**0.8981**	0.248	27.6	ug/L	42	Standard
>	**Ga**	71	78.3	22.4				mg/L	42	Standard
	Rb	85	147149.5	0.7				ug/L	17	Standard
	Y	89	405581.1	1.8				ug/L	417098	Standard
>	**Rh**	103	178.3	29.5				ug/L	13	Standard
	Mo	98	220.6	12.3	**0.0817**	0.013	15.7	ug/L	11	Standard
	Ag	107	177.0	68.0	**0.0160**	0.020	127.1	ug/L	68	Standard
	Cd	111	65.9	51.9	**0.0256**	0.018	71.0	mg/L	2	Standard
	Cd	114	151.1	49.7	**0.0291**	0.016	55.4	ug/L	12	Standard
>	**In**	115	557347.0	2.4				ug/L	587441	Standard
	Sn	118	766.7	0.4	**0.0437**	0.004	8.6	ug/L	497	Standard
	Sb	123	305.8	17.0	**0.0650**	0.013	20.6	ug/L	78	Standard
	Ba	135	5047.2	3.4	**2.5853**	0.149	5.7	ug/L	24	Standard
	Ce	140	491.7	100.3				ug/L	25	Standard
>	**Tb**	159	935324.5	2.7				ug/L	939701	Standard
	Ho	165	18.3	103.3				ug/L	10	Standard
	Tl	203	947.0	9.8	**0.1179**	0.014	11.5	ug/L	4	Standard
	Tl	205	846.7	27.0	**0.1237**	0.035	28.6	ug/L	2	Standard
	Pb	206	418.3	37.6	**0.0106**	0.032	302.9	ug/L	240	Standard
	Pb	207	351.7	32.0	**0.0074**	0.026	354.3	ug/L	198	Standard
	Pb	208	1404.0	30.0	**0.0041**	0.025	608.6	ug/L	883	Standard
	U	238	972.4	9.3	**0.1427**	0.016	11.1	ug/L	3	Standard
>	**Bi**	209	503603.8	1.5				ug/L	523984	Standard

Sample ID: L1606124702DP WG574001-06
Report Date/Time: Wednesday, June 29, 2016 16:00:44
Page 1

Approved: July 04, 2016

Na	23	25.0	34.6	**311.2656**	112.939	36.3	mg/L	0	Standard
Mg	24	85343.0	0.9	**118.3682**	4.384	3.7	mg/L	37	Standard
K	39	1298.4	2.9	**2.9712**	0.147	4.9	mg/L	38	Standard
Ca	43	2830.3	7.5	**189.4839**	12.043	6.4	mg/L	10	Standard
Fe	54	146.6	14.0	**-0.0044**	0.021	483.2	mg/L	179	Standard
Fe	57	1698.4	7.8	**5.1376**	0.623	12.1	mg/L	335	Standard
Sc-1	45	56811.4	3.1				mg/L	61029	Standard
Cl	35	65999.5	1.4				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	22520.1	6.6				ug/L	1547	Standard
P	31	16799.3	3.7				ug/L	15286	Standard
S	34	2040.1	3.0				ug/L	1572	Standard
Sr	88	920.0	8.9				ug/L	73	Standard
C	12	83.3	25.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	173.2				mg/L	3	Standard
Dy	164	32.1	98.6				mg/L	6	Standard
Ho-1	165	18.3	103.3				mg/L	10	Standard
Er	166	26.7	43.3				mg/L	13	Standard
I	127	45859.6	7.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		203.086	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		95.605	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606124702DP WG574001-06
Report Date/Time: Wednesday, June 29, 2016 16:00:44
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.877
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.110
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124702DP WG574001-06
Report Date/Time: Wednesday, June 29, 2016 16:00:44
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126501 WG574001-02
Sample Date/Time: Wednesday, June 29, 2016 16:01:39
Number of Replicates: 3
Autosampler Position: 209
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	112017.9	3.5				ug/L	74699	Standard
	Be	9	21.7	35.3	**0.0242**	0.006	26.3	ug/L	8	Standard
	Al	27	58851331.1	1.0	**267.2037**	7.701	2.9	ug/L	1493	Standard
	Sc	45	63265.2	2.1				ug/L	61029	Standard
	Ti	47	314.3	6.0	**0.7250**	0.050	6.9	ug/L	35	Standard
	V	51	84379.6	0.6	**11.0401**	0.098	0.9	ug/L	1435	Standard
	Cr	52	25344.2	1.7	**2.4155**	0.065	2.7	ug/L	7635	Standard
	Cr	53	30644.3	3.7	**33.0161**	1.295	3.9	ug/L	363	Standard
	Mn	55	3058.3	0.3	**0.2937**	0.003	0.9	ug/L	747	Standard
	Co	59	981.4	1.2	**0.1122**	0.002	1.4	ug/L	234	Standard
	Ni	60	4598.7	1.8	**2.3463**	0.051	2.2	ug/L	215	Standard
	Cu	65	1226.7	2.1	**0.5592**	0.015	2.8	ug/L	104	Standard
	Zn	66	2978.0	3.2	**2.2873**	0.091	4.0	ug/L	363	Standard
>	**Ge**	72	492356.2	0.3				ug/L	495142	Standard
	As	75	3467.1	5.3	**3.1728**	0.175	5.5	ug/L	-50	Standard
	Se	82	1561.8	3.2	**14.1732**	0.483	3.4	ug/L	14	Standard
	Se-1	77	3085.3	7.4	**41.8633**	3.228	7.7	ug/L	42	Standard
>	**Ga**	71	63.3	27.7				mg/L	42	Standard
	Rb	85	9206.1	3.0				ug/L	17	Standard
	Y	89	442169.1	3.5				ug/L	417098	Standard
>	**Rh**	103	255.0	11.9				ug/L	13	Standard
	Mo	98	5205.2	3.3	**1.8993**	0.033	1.7	ug/L	11	Standard
	Ag	107	803.7	5.3	**0.1074**	0.005	4.8	ug/L	68	Standard
	Cd	111	45.0	9.2	**0.0131**	0.002	16.6	mg/L	2	Standard
	Cd	114	163.3	22.8	**0.0289**	0.007	22.9	ug/L	12	Standard
>	**In**	115	602229.8	3.0				ug/L	587441	Standard
	Sn	118	683.3	9.8	**0.0193**	0.009	45.4	ug/L	497	Standard
	Sb	123	210.7	19.8	**0.0397**	0.008	19.9	ug/L	78	Standard
	Ba	135	21407.4	0.7	**10.1914**	0.253	2.5	ug/L	24	Standard
	Ce	140	328.3	10.1				ug/L	25	Standard
>	**Tb**	159	984916.0	0.2				ug/L	939701	Standard
	Ho	165	35.0	24.7				ug/L	10	Standard
	Tl	203	94.3	9.0	**0.0098**	0.001	12.0	ug/L	4	Standard
	Tl	205	65.0	40.7	**0.0094**	0.004	40.6	ug/L	2	Standard
	Pb	206	401.3	3.0	**0.0049**	0.001	28.2	ug/L	240	Standard
	Pb	207	354.0	5.2	**0.0056**	0.003	55.2	ug/L	198	Standard
	Pb	208	1436.0	3.5	**0.0035**	0.003	74.9	ug/L	883	Standard
	U	238	31993.4	1.5	**4.7067**	0.092	1.9	ug/L	3	Standard
>	**Bi**	209	517006.9	1.2				ug/L	523984	Standard

Sample ID: L1606126501 WG574001-02
Report Date/Time: Wednesday, June 29, 2016 16:03:56
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**55.9813**	56.756	101.4	mg/L	0	Standard
Mg	24	180057.1	1.6	**224.1644**	1.858	0.8	mg/L	37	Standard
K	39	641.7	5.8	**1.2549**	0.096	7.6	mg/L	38	Standard
Ca	43	708.3	7.0	**40.6474**	3.060	7.5	mg/L	10	Standard
Fe	54	157.9	5.8	**-0.0091**	0.008	86.4	mg/L	179	Standard
Fe	57	550.0	17.3	**0.7314**	0.333	45.5	mg/L	335	Standard
> **Sc-1**	45	63265.2	2.1				mg/L	61029	Standard
Cl	35	33676.1	1.7				ug/L	20195	Standard
Kr	83	0.3	173.2				ug/L	4	Standard
Br	81	48808.6	5.5				ug/L	1547	Standard
P	31	23084.3	3.5				ug/L	15286	Standard
S	34	2106.8	4.3				ug/L	1572	Standard
Sr	88	1195.0	6.7				ug/L	73	Standard
C	12	130.0	40.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	9.0	113.3				mg/L	6	Standard
Ho-1	165	35.0	24.7				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	49504.2	3.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		149.959	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.437	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126501 WG574001-02
Report Date/Time: Wednesday, June 29, 2016 16:03:56
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.517
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.668
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606126501 WG574001-02
Report Date/Time: Wednesday, June 29, 2016 16:03:56
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126502S WG574001-07
Sample Date/Time: Wednesday, June 29, 2016 16:04:50
Number of Replicates: 3
Autosampler Position: 210
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	109865.8	4.5				ug/L	74699	Standard
	Be	9	61395.8	1.0	**47.4912**	2.348	4.9	ug/L	8	Standard
	Al	27	61297739.1	1.7	**284.0961**	17.361	6.1	ug/L	1493	Standard
	Sc	45	64222.6	2.9				ug/L	61029	Standard
	Ti	47	334.7	6.1	**0.7462**	0.065	8.7	ug/L	35	Standard
	V	51	514269.3	0.9	**65.7128**	0.488	0.7	ug/L	1435	Standard
	Cr	52	414721.1	1.2	**53.1400**	0.316	0.6	ug/L	7635	Standard
	Cr	53	79122.8	2.2	**82.5246**	0.802	1.0	ug/L	363	Standard
	Mn	55	379075.2	0.4	**52.5566**	0.801	1.5	ug/L	747	Standard
	Co	59	375122.1	1.7	**54.9533**	1.379	2.5	ug/L	234	Standard
	Ni	60	98886.4	1.8	**50.9941**	0.435	0.9	ug/L	215	Standard
	Cu	65	94781.9	0.9	**49.1885**	0.428	0.9	ug/L	104	Standard
	Zn	66	59708.1	1.0	**52.4531**	0.349	0.7	ug/L	363	Standard
>	Ge	72	511580.8	1.5				ug/L	495142	Standard
	As	75	64226.1	0.6	**55.5617**	0.582	1.0	ug/L	-50	Standard
	Se	82	7637.7	1.0	**67.2251**	1.117	1.7	ug/L	14	Standard
	Se-1	77	7177.4	2.9	**94.4764**	1.329	1.4	ug/L	42	Standard
>	Ga	71	96.7	7.9				mg/L	42	Standard
	Rb	85	9571.4	0.6				ug/L	17	Standard
	Y	89	459795.2	1.7				ug/L	417098	Standard
>	Rh	103	293.3	12.0				ug/L	13	Standard
	Mo	98	5421.1	3.7	**1.9704**	0.066	3.4	ug/L	11	Standard
	Ag	107	318338.0	1.0	**47.4485**	0.356	0.8	ug/L	68	Standard
	Cd	111	104991.2	1.0	**48.7852**	0.464	1.0	mg/L	2	Standard
	Cd	114	250399.7	1.5	**47.4405**	0.760	1.6	ug/L	12	Standard
>	In	115	604563.1	0.3				ug/L	587441	Standard
	Sn	118	798.4	4.4	**0.0381**	0.006	16.0	ug/L	497	Standard
	Sb	123	240747.3	1.0	**50.1002**	0.397	0.8	ug/L	78	Standard
	Ba	135	122571.5	0.3	**58.1746**	0.379	0.7	ug/L	24	Standard
	Ce	140	328.3	0.9				ug/L	25	Standard
>	Tb	159	981221.4	1.4				ug/L	939701	Standard
	Ho	165	28.3	44.4				ug/L	10	Standard
	Tl	203	414235.7	0.7	**51.0974**	0.645	1.3	ug/L	4	Standard
	Tl	205	378219.2	0.6	**53.7534**	0.090	0.2	ug/L	2	Standard
	Pb	206	277784.6	0.1	**52.8053**	0.305	0.6	ug/L	240	Standard
	Pb	207	237018.9	0.5	**50.7105**	0.514	1.0	ug/L	198	Standard
	Pb	208	965977.0	0.4	**52.3096**	0.506	1.0	ug/L	883	Standard
	U	238	407586.9	0.7	**60.2072**	0.700	1.2	ug/L	3	Standard
>	Bi	209	515321.0	0.6				ug/L	523984	Standard

Sample ID: L1606126502S WG574001-07
Report Date/Time: Wednesday, June 29, 2016 16:07:07
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.6720**	30.600	173.2	mg/L	0	Standard
Mg	24	189285.1	1.3	**232.3057**	9.548	4.1	mg/L	37	Standard
K	39	641.7	6.1	**1.2328**	0.045	3.6	mg/L	38	Standard
Ca	43	760.0	12.3	**43.0240**	4.391	10.2	mg/L	10	Standard
Fe	54	185.4	16.8	**0.0114**	0.028	245.2	mg/L	179	Standard
Fe	57	630.0	4.2	**0.9612**	0.087	9.0	mg/L	335	Standard
> **Sc-1**	45	64222.6	2.9				mg/L	61029	Standard
Cl	35	34933.6	1.5				ug/L	20195	Standard
Kr	83	2.7	114.6				ug/L	4	Standard
Br	81	53306.9	1.0				ug/L	1547	Standard
P	31	23740.3	1.4				ug/L	15286	Standard
S	34	2118.5	3.3				ug/L	1572	Standard
Sr	88	1145.0	3.3				ug/L	73	Standard
C	12	140.0	21.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	28.6	38.4				mg/L	6	Standard
Ho-1	165	28.3	44.4				mg/L	10	Standard
Er	166	30.0	66.7				mg/L	13	Standard
I	127	81496.0	7.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		147.078	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.320	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126502S WG574001-07
Report Date/Time: Wednesday, June 29, 2016 16:07:07
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	102.915	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	98.347	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606126502S WG574001-07
Report Date/Time: Wednesday, June 29, 2016 16:07:07
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126503SD WG574001-08
Sample Date/Time: Wednesday, June 29, 2016 16:08:02
Number of Replicates: 3
Autosampler Position: 211
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	108017.3	3.5				ug/L	74699	Standard
	Be	9	61104.9	5.2	**47.9870**	0.846	1.8	ug/L	8	Standard
	Al	27	60436124.3	2.1	**284.5445**	9.073	3.2	ug/L	1493	Standard
	Sc	45	65195.1	3.6				ug/L	61029	Standard
	Ti	47	321.3	5.9	**0.6738**	0.054	8.0	ug/L	35	Standard
	V	51	521782.6	1.8	**63.5667**	1.206	1.9	ug/L	1435	Standard
	Cr	52	431117.5	0.3	**52.6702**	0.983	1.9	ug/L	7635	Standard
	Cr	53	81998.0	1.1	**81.5509**	0.729	0.9	ug/L	363	Standard
	Mn	55	379421.8	1.7	**50.1647**	1.838	3.7	ug/L	747	Standard
	Co	59	378746.6	2.0	**52.9167**	2.098	4.0	ug/L	234	Standard
	Ni	60	101643.5	0.7	**49.9872**	1.144	2.3	ug/L	215	Standard
	Cu	65	96036.4	1.2	**47.5169**	0.455	1.0	ug/L	104	Standard
	Zn	66	60629.4	1.5	**50.7689**	0.479	0.9	ug/L	363	Standard
>	**Ge**	72	536566.5	1.9				ug/L	495142	Standard
	As	75	65668.7	0.7	**54.1701**	0.796	1.5	ug/L	-50	Standard
	Se	82	7769.6	1.1	**65.2136**	1.970	3.0	ug/L	14	Standard
	Se-1	77	7316.4	1.6	**91.8340**	1.824	2.0	ug/L	42	Standard
>	**Ga**	71	56.7	22.2				mg/L	42	Standard
	Rb	85	9684.8	3.9				ug/L	17	Standard
	Y	89	458590.9	3.1				ug/L	417098	Standard
>	**Rh**	103	275.0	12.6				ug/L	13	Standard
	Mo	98	5397.9	1.9	**1.9792**	0.040	2.0	ug/L	11	Standard
	Ag	107	315299.7	0.9	**47.4082**	1.008	2.1	ug/L	68	Standard
	Cd	111	105649.0	0.5	**49.5183**	0.751	1.5	mg/L	2	Standard
	Cd	114	256574.5	1.5	**49.0334**	1.028	2.1	ug/L	12	Standard
>	**In**	115	599465.1	2.0				ug/L	587441	Standard
	Sn	118	820.0	11.4	**0.0428**	0.015	34.4	ug/L	497	Standard
	Sb	123	245153.5	0.2	**51.4644**	0.950	1.8	ug/L	78	Standard
	Ba	135	122768.4	0.8	**58.7829**	1.524	2.6	ug/L	24	Standard
	Ce	140	333.3	10.5				ug/L	25	Standard
>	**Tb**	159	983758.7	0.8				ug/L	939701	Standard
	Ho	165	18.3	41.7				ug/L	10	Standard
	Tl	203	416470.0	0.7	**50.8907**	0.470	0.9	ug/L	4	Standard
	Tl	205	377588.7	0.1	**53.1619**	0.172	0.3	ug/L	2	Standard
	Pb	206	279930.3	0.5	**52.7144**	0.351	0.7	ug/L	240	Standard
	Pb	207	238634.3	0.7	**50.5762**	0.275	0.5	ug/L	198	Standard
	Pb	208	958199.8	1.0	**51.4011**	0.688	1.3	ug/L	883	Standard
	U	238	405052.0	0.8	**59.2716**	0.671	1.1	ug/L	3	Standard
>	**Bi**	209	520189.0	0.3				ug/L	523984	Standard

Sample ID: L1606126503SD WG574001-08
Report Date/Time: Wednesday, June 29, 2016 16:10:18
Page 1

Approved: July 04, 2016

Na	23	5.0	173.2	**51.9909**	90.042	173.2	mg/L	0	Standard
Mg	24	184795.4	2.1	**223.5320**	12.550	5.6	mg/L	37	Standard
K	39	668.3	13.7	**1.2668**	0.164	13.0	mg/L	38	Standard
Ca	43	725.0	7.3	**40.3567**	2.975	7.4	mg/L	10	Standard
Fe	54	151.4	21.3	**-0.0183**	0.024	132.3	mg/L	179	Standard
Fe	57	623.3	16.5	**0.9045**	0.268	29.6	mg/L	335	Standard
> **Sc-1**	45	65195.1	3.6				mg/L	61029	Standard
Cl	35	33686.7	1.5				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	51370.3	3.2				ug/L	1547	Standard
P	31	23124.3	1.4				ug/L	15286	Standard
S	34	2101.8	2.3				ug/L	1572	Standard
Sr	88	1251.7	9.6				ug/L	73	Standard
C	12	163.3	19.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	9.4	104.3				mg/L	6	Standard
Ho-1	165	18.3	41.7				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	90519.1	8.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		144.604	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.366	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126503SD WG574001-08
Report Date/Time: Wednesday, June 29, 2016 16:10:18
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	102.047	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	99.276	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Sample ID: L1606126503SD WG574001-08
Report Date/Time: Wednesday, June 29, 2016 16:10:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701
Sample Date/Time: Wednesday, June 29, 2016 16:11:13
Number of Replicates: 3
Autosampler Position: 212
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	160464.4	3.8				ug/L	74699	Standard
	Be	9	12485.1	1.8	**6.6126**	0.132	2.0	ug/L	8	Standard
	Al	27	129371635.5	1.7	**410.0067**	8.683	2.1	ug/L	1493	Standard
	Sc	45	97873.5	2.0				ug/L	61029	Standard
	Ti	47	12940.2	2.9	**32.4229**	1.439	4.4	ug/L	35	Standard
	V	51	62587.9	0.8	**7.6988**	0.173	2.3	ug/L	1435	Standard
	Cr	52	130472.4	1.2	**15.7449**	0.176	1.1	ug/L	7635	Standard
	Cr	53	17299.9	2.8	**17.4949**	0.102	0.6	ug/L	363	Standard
	Mn	55	53885293.3	1.0	**7372.4056**	142.217	1.9	ug/L	747	Standard
	Co	59	1486809.2	2.0	**214.4609**	6.754	3.1	ug/L	234	Standard
	Ni	60	953927.1	1.9	**485.1518**	7.673	1.6	ug/L	215	Standard
	Cu	65	14021.8	2.7	**7.0705**	0.048	0.7	ug/L	104	Standard
	Zn	66	1250022.8	1.4	**1089.5042**	11.046	1.0	ug/L	363	Standard
>	**Ge**	72	519921.5	2.2				ug/L	495142	Standard
	As	75	2369.9	1.5	**2.0748**	0.031	1.5	ug/L	-50	Standard
	Se	82	334.2	2.8	**2.7639**	0.140	5.1	ug/L	14	Standard
	Se-1	77	477.7	9.4	**5.6016**	0.492	8.8	ug/L	42	Standard
>	**Ga**	71	3503.7	1.4				mg/L	42	Standard
	Rb	85	155749.9	1.6				ug/L	17	Standard
	Y	89	3051185.4	0.7				ug/L	417098	Standard
>	**Rh**	103	200.0	9.0				ug/L	13	Standard
	Mo	98	648.1	0.8	**0.2277**	0.002	1.0	ug/L	11	Standard
	Ag	107	406.3	10.5	**0.0469**	0.006	13.8	ug/L	68	Standard
	Cd	111	4037.2	2.2	**1.8458**	0.056	3.0	mg/L	2	Standard
	Cd	114	10321.7	4.5	**1.9292**	0.065	3.4	ug/L	12	Standard
>	**In**	115	612067.4	1.8				ug/L	587441	Standard
	Sn	118	936.7	6.3	**0.0593**	0.012	20.5	ug/L	497	Standard
	Sb	123	674.3	5.8	**0.1344**	0.006	4.1	ug/L	78	Standard
	Ba	135	15717.8	1.3	**7.3556**	0.136	1.8	ug/L	24	Standard
	Ce	140	2639736.8	0.7				ug/L	25	Standard
>	**Tb**	159	1213429.7	1.8				ug/L	939701	Standard
	Ho	165	212831.1	0.7				ug/L	10	Standard
	Tl	203	2423.5	3.4	**0.2853**	0.011	3.9	ug/L	4	Standard
	Tl	205	2225.2	6.3	**0.3038**	0.017	5.8	ug/L	2	Standard
	Pb	206	7086.3	0.1	**1.2240**	0.010	0.8	ug/L	240	Standard
	Pb	207	5547.0	2.2	**1.0710**	0.022	2.1	ug/L	198	Standard
	Pb	208	23218.6	1.0	**1.1348**	0.005	0.5	ug/L	883	Standard
	U	238	25740.2	1.8	**3.6463**	0.044	1.2	ug/L	3	Standard
>	**Bi**	209	536711.9	0.7				ug/L	523984	Standard

Sample ID: L1606124701
Report Date/Time: Wednesday, June 29, 2016 16:13:30
Page 1

Approved: July 04, 2016

Na	23	40.0	57.3	**285.3574**	160.270	56.2	mg/L	0	Standard
Mg	24	87773.6	1.9	**70.5914**	1.090	1.5	mg/L	37	Standard
K	39	1211.7	2.8	**1.5562**	0.053	3.4	mg/L	38	Standard
Ca	43	3308.7	4.2	**127.7939**	4.640	3.6	mg/L	10	Standard
Fe	54	180981.2	2.7	**95.5321**	0.892	0.9	mg/L	179	Standard
Fe	57	48873.8	3.4	**102.4969**	2.098	2.0	mg/L	335	Standard
Sc-1	45	97873.5	2.0				mg/L	61029	Standard
Cl	35	64420.7	1.3				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	26873.8	3.7				ug/L	1547	Standard
P	31	52208.1	1.4				ug/L	15286	Standard
S	34	1956.8	2.0				ug/L	1572	Standard
Sr	88	1171.7	10.6				ug/L	73	Standard
C	12	126.7	46.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	329920.1	1.3				mg/L	6	Standard
Ho-1	165	212831.1	0.7				mg/L	10	Standard
Er	166	185158.1	1.8				mg/L	13	Standard
I	127	88143.1	6.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		214.815	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		105.004	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606124701
Report Date/Time: Wednesday, June 29, 2016 16:13:30
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.192
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.429
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701
Report Date/Time: Wednesday, June 29, 2016 16:13:30
Page 3

Approved: July 04, 2016

Co 59 Upper, S, EEE	Co	59
Ni 60 Upper, S, EEE	Ni	60
Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701PS WG574688-01
Sample Date/Time: Wednesday, June 29, 2016 16:14:25
Number of Replicates: 3
Autosampler Position: 213
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	166369.0	3.1				ug/L	74699	Standard
	Be	9	78228.1	1.5	**39.9183**	0.744	1.9	ug/L	8	Standard
	Al	27	127858501.0	1.9	**390.7420**	6.863	1.8	ug/L	1493	Standard
	Sc	45	95075.8	2.0				ug/L	61029	Standard
	Ti	47	12076.5	3.4	**30.6847**	0.840	2.7	ug/L	35	Standard
	V	51	487373.4	0.7	**62.1823**	0.880	1.4	ug/L	1435	Standard
	Cr	52	506749.0	0.5	**65.0704**	0.602	0.9	ug/L	7635	Standard
	Cr	53	63546.3	1.7	**66.1298**	1.201	1.8	ug/L	363	Standard
	Mn	55	53730576.4	0.3	**7459.5309**	74.503	1.0	ug/L	747	Standard
	Co	59	1866113.9	0.2	**273.1251**	3.396	1.2	ug/L	234	Standard
	Ni	60	1009610.1	0.8	**521.0961**	4.502	0.9	ug/L	215	Standard
	Cu	65	102190.2	1.2	**52.9646**	0.037	0.1	ug/L	104	Standard
	Zn	66	1265218.7	0.3	**1119.1475**	11.121	1.0	ug/L	363	Standard
>	**Ge**	72	512284.2	1.2				ug/L	495142	Standard
	As	75	65072.9	0.3	**56.2148**	0.511	0.9	ug/L	-50	Standard
	Se	82	7179.2	1.0	**63.0878**	0.484	0.8	ug/L	14	Standard
	Se-1	77	4836.8	0.7	**63.3901**	0.749	1.2	ug/L	42	Standard
>	**Ga**	71	3398.7	7.2				mg/L	42	Standard
	Rb	85	158551.9	0.5				ug/L	17	Standard
	Y	89	3079691.2	1.3				ug/L	417098	Standard
>	**Rh**	103	243.3	6.3				ug/L	13	Standard
	Mo	98	610.5	4.6	**0.2160**	0.008	3.7	ug/L	11	Standard
	Ag	107	329069.5	0.6	**48.8691**	0.577	1.2	ug/L	68	Standard
	Cd	111	109320.3	0.2	**50.6101**	0.459	0.9	mg/L	2	Standard
	Cd	114	265491.9	0.4	**50.1152**	0.608	1.2	ug/L	12	Standard
>	**In**	115	606831.5	1.0				ug/L	587441	Standard
	Sn	118	1098.4	5.9	**0.0874**	0.012	14.0	ug/L	497	Standard
	Sb	123	240923.7	1.8	**49.9558**	1.171	2.3	ug/L	78	Standard
	Ba	135	113288.6	0.8	**53.5705**	0.782	1.5	ug/L	24	Standard
	Ce	140	2588807.2	1.1				ug/L	25	Standard
>	**Tb**	159	1203377.0	1.6				ug/L	939701	Standard
	Ho	165	210409.6	1.8				ug/L	10	Standard
	Tl	203	437720.2	1.4	**52.3305**	0.744	1.4	ug/L	4	Standard
	Tl	205	404409.4	0.6	**55.7069**	0.222	0.4	ug/L	2	Standard
	Pb	206	295264.9	0.3	**54.4023**	0.212	0.4	ug/L	240	Standard
	Pb	207	250626.5	0.1	**51.9726**	0.335	0.6	ug/L	198	Standard
	Pb	208	1033784.3	0.4	**54.2614**	0.547	1.0	ug/L	883	Standard
	U	238	423217.8	0.6	**60.5919**	0.666	1.1	ug/L	3	Standard
>	**Bi**	209	531687.2	0.6				ug/L	523984	Standard

Sample ID: L1606124701PS WG574688-01
Report Date/Time: Wednesday, June 29, 2016 16:16:41
Page 1

Approved: July 04, 2016

Na	23	31.7	32.9	**235.3667**	80.946	34.4	mg/L	0	Standard
Mg	24	88559.7	0.9	**73.3333**	1.498	2.0	mg/L	37	Standard
K	39	1311.7	6.5	**1.7460**	0.092	5.3	mg/L	38	Standard
Ca	43	3437.1	1.4	**136.8797**	3.742	2.7	mg/L	10	Standard
Fe	54	183063.9	0.7	**99.5077**	1.370	1.4	mg/L	179	Standard
Fe	57	49900.4	2.7	**107.8438**	4.358	4.0	mg/L	335	Standard
Sc-1	45	95075.8	2.0				mg/L	61029	Standard
Cl	35	63889.8	1.2				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	26186.0	7.9				ug/L	1547	Standard
P	31	52633.0	2.3				ug/L	15286	Standard
S	34	1903.5	2.7				ug/L	1572	Standard
Sr	88	1258.4	8.1				ug/L	73	Standard
C	12	153.3	16.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	329580.7	1.9				mg/L	6	Standard
Ho-1	165	210409.6	1.8				mg/L	10	Standard
Er	166	182077.6	1.3				mg/L	13	Standard
I	127	87547.6	5.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		222.720	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		103.462	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606124701PS WG574688-01
Report Date/Time: Wednesday, June 29, 2016 16:16:41
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	103.301
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	101.470
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701PS WG574688-01
Report Date/Time: Wednesday, June 29, 2016 16:16:41
Page 3

Approved: July 04, 2016

Co 59 Upper, S, EEE	Co	59
Ni 60 Upper, S, EEE	Ni	60
Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701SDL WG574688-02
Sample Date/Time: Wednesday, June 29, 2016 16:17:36
Number of Replicates: 3
Autosampler Position: 214
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	115043.6	2.7				ug/L	74699	Standard
	Be	9	2510.2	6.0	**1.8579**	0.064	3.4	ug/L	8	Standard
	Al	27	26603666.2	1.3	**117.5773**	3.118	2.7	ug/L	1493	Standard
	Sc	45	70520.4	2.0				ug/L	61029	Standard
	Ti	47	2655.2	2.4	**6.2496**	0.159	2.5	ug/L	35	Standard
	V	51	13651.6	1.3	**1.4451**	0.028	2.0	ug/L	1435	Standard
	Cr	52	32029.5	0.6	**2.9000**	0.113	3.9	ug/L	7635	Standard
	Cr	53	4062.2	6.0	**3.6565**	0.147	4.0	ug/L	363	Standard
	Mn	55	11326454.5	1.3	**1476.9710**	52.700	3.6	ug/L	747	Standard
	Co	59	315829.2	0.8	**43.3845**	1.293	3.0	ug/L	234	Standard
	Ni	60	206220.6	0.5	**99.8677**	2.554	2.6	ug/L	215	Standard
	Cu	65	3227.7	3.3	**1.4714**	0.087	5.9	ug/L	104	Standard
	Zn	66	276031.8	0.8	**228.9686**	6.765	3.0	ug/L	363	Standard
>	**Ge**	72	545622.5	2.3				ug/L	495142	Standard
	As	75	529.0	5.1	**0.4886**	0.031	6.3	ug/L	-50	Standard
	Se	82	109.8	7.8	**0.7692**	0.061	7.9	ug/L	14	Standard
	Se-1	77	127.7	10.6	**0.9669**	0.203	21.0	ug/L	42	Standard
>	**Ga**	71	896.7	4.3				mg/L	42	Standard
	Rb	85	33246.4	0.8				ug/L	17	Standard
	Y	89	1035711.2	2.7				ug/L	417098	Standard
>	**Rh**	103	68.3	33.0				ug/L	13	Standard
	Mo	98	137.2	4.4	**0.0416**	0.001	3.0	ug/L	11	Standard
	Ag	107	193.7	13.0	**0.0144**	0.003	20.5	ug/L	68	Standard
	Cd	111	922.1	1.5	**0.3978**	0.011	2.7	mg/L	2	Standard
	Cd	114	2202.4	8.0	**0.3929**	0.032	8.2	ug/L	12	Standard
>	**In**	115	638835.9	2.4				ug/L	587441	Standard
	Sn	118	601.7	4.7	**-0.0000**	0.004	76937.0	ug/L	497	Standard
	Sb	123	1960.6	22.3	**0.3834**	0.095	24.7	ug/L	78	Standard
	Ba	135	3410.7	1.0	**1.5171**	0.049	3.2	ug/L	24	Standard
	Ce	140	553030.9	1.0				ug/L	25	Standard
>	**Tb**	159	1064988.1	1.7				ug/L	939701	Standard
	Ho	165	44299.7	1.5				ug/L	10	Standard
	Tl	203	596.3	8.1	**0.0662**	0.006	8.4	ug/L	4	Standard
	Tl	205	613.3	10.1	**0.0808**	0.008	10.1	ug/L	2	Standard
	Pb	206	1759.1	0.5	**0.2383**	0.002	0.8	ug/L	240	Standard
	Pb	207	1407.7	5.3	**0.2088**	0.015	7.2	ug/L	198	Standard
	Pb	208	5898.4	2.6	**0.2216**	0.008	3.7	ug/L	883	Standard
	U	238	5357.3	1.2	**0.7271**	0.010	1.3	ug/L	3	Standard
>	**Bi**	209	557502.4	0.2				ug/L	523984	Standard

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:19:53
Page 1

Approved: July 04, 2016

Na	23	11.7	89.2	**117.5100**	105.501	89.8	mg/L	0	Standard
Mg	24	18105.8	1.6	**20.1624**	0.144	0.7	mg/L	37	Standard
K	39	321.7	3.6	**0.5016**	0.035	6.9	mg/L	38	Standard
Ca	43	733.4	8.9	**37.5359**	2.964	7.9	mg/L	10	Standard
Fe	54	38163.2	2.6	**27.8607**	0.186	0.7	mg/L	179	Standard
Fe	57	10221.8	1.3	**28.9959**	0.316	1.1	mg/L	335	Standard
> **Sc-1**	45	70520.4	2.0				mg/L	61029	Standard
Cl	35	31180.1	1.1				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	7325.1	3.6				ug/L	1547	Standard
P	31	24873.7	0.5				ug/L	15286	Standard
S	34	1646.8	8.5				ug/L	1572	Standard
Sr	88	296.7	11.0				ug/L	73	Standard
C	12	96.7	53.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	69181.7	2.0				mg/L	6	Standard
Ho-1	165	44299.7	1.5				mg/L	10	Standard
Er	166	39063.6	1.3				mg/L	13	Standard
I	127	21375.1	3.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		154.010	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.195	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:19:53
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.749
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.397
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:19:53
Page 3

Approved: July 04, 2016

Zn 66 Upper, S, EEE	Zn	66

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 16:20:49
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	104492.8	3.9				ug/L	74699	Standard
	Be	9	63869.3	1.0	**51.9084**	1.516	2.9	ug/L	8	Standard
	Al	27	9148865.9	2.6	**44.5067**	0.724	1.6	ug/L	1493	Standard
	Sc	45	68447.7	3.2				ug/L	61029	Standard
	Ti	47	43923.6	1.6	**106.2047**	1.334	1.3	ug/L	35	Standard
	V	51	454775.2	0.5	**55.0469**	0.468	0.8	ug/L	1435	Standard
	Cr	52	431898.3	0.6	**52.4410**	0.316	0.6	ug/L	7635	Standard
	Cr	53	53343.8	2.8	**52.6203**	1.459	2.8	ug/L	363	Standard
	Mn	55	427711.0	4.2	**56.2006**	1.985	3.5	ug/L	747	Standard
	Co	59	425402.4	1.1	**59.0619**	0.614	1.0	ug/L	234	Standard
	Ni	60	106741.4	0.9	**52.1773**	0.185	0.4	ug/L	215	Standard
	Cu	65	106457.4	0.4	**52.3703**	0.176	0.3	ug/L	104	Standard
	Zn	66	63607.9	0.7	**52.9682**	0.530	1.0	ug/L	363	Standard
>	Ge	72	539725.1	0.7				ug/L	495142	Standard
	As	75	63945.9	1.5	**52.4378**	1.123	2.1	ug/L	-50	Standard
	Se	82	6584.0	0.6	**54.8988**	0.684	1.2	ug/L	14	Standard
	Se-1	77	4386.3	1.4	**54.4772**	1.111	2.0	ug/L	42	Standard
>	Ga	71	80.0	16.5				mg/L	42	Standard
	Rb	85	970.0	8.8				ug/L	17	Standard
	Y	89	494238.2	2.5				ug/L	417098	Standard
>	Rh	103	45.0	11.1				ug/L	13	Standard
	Mo	98	317512.9	0.8	**102.4795**	1.834	1.8	ug/L	11	Standard
	Ag	107	378789.7	0.9	**49.9860**	0.591	1.2	ug/L	68	Standard
	Cd	111	119686.6	0.4	**49.2403**	0.865	1.8	mg/L	2	Standard
	Cd	114	291890.5	1.4	**48.9669**	1.322	2.7	ug/L	12	Standard
>	In	115	682927.0	1.5				ug/L	587441	Standard
	Sn	118	338721.1	0.8	**49.7884**	0.723	1.5	ug/L	497	Standard
	Sb	123	263759.6	1.2	**48.5931**	0.331	0.7	ug/L	78	Standard
	Ba	135	111614.3	0.7	**46.9010**	0.972	2.1	ug/L	24	Standard
	Ce	140	1265.1	98.5				ug/L	25	Standard
>	Tb	159	1064058.1	1.0				ug/L	939701	Standard
	Ho	165	88.3	85.2				ug/L	10	Standard
	Tl	203	475263.7	0.7	**51.4258**	1.211	2.4	ug/L	4	Standard
	Tl	205	438160.3	0.4	**54.6297**	1.400	2.6	ug/L	2	Standard
	Pb	206	306629.0	0.6	**51.1326**	1.396	2.7	ug/L	240	Standard
	Pb	207	272793.9	0.3	**51.2040**	1.495	2.9	ug/L	198	Standard
	Pb	208	1100211.2	0.6	**52.2716**	1.709	3.3	ug/L	883	Standard
	U	238	414469.5	0.5	**53.7134**	1.693	3.2	ug/L	3	Standard
>	Bi	209	587705.0	2.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 16:23:06
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**35.5299**	61.531	173.2	mg/L	0	Standard
Mg	24	4809.1	8.0	**5.4678**	0.353	6.5	mg/L	37	Standard
K	39	2090.1	1.6	**4.0082**	0.167	4.2	mg/L	38	Standard
Ca	43	111.7	10.3	**3.7837**	0.862	22.8	mg/L	10	Standard
Fe	54	7082.4	2.6	**5.2191**	0.217	4.1	mg/L	179	Standard
Fe	57	2040.1	4.3	**5.1108**	0.267	5.2	mg/L	335	Standard
Sc-1	45	68447.7	3.2				mg/L	61029	Standard
Cl	35	25830.7	1.4				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	2190.2	6.5				ug/L	1547	Standard
P	31	18406.2	0.9				ug/L	15286	Standard
S	34	1965.1	6.7				ug/L	1572	Standard
Sr	88	71.7	21.3				ug/L	73	Standard
C	12	103.3	29.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	152.5	102.6				mg/L	6	Standard
Ho-1	165	88.3	85.2				mg/L	10	Standard
Er	166	86.7	74.2				mg/L	13	Standard
I	127	2766.9	21.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	103.817		
Al	27	89.013		
Sc	45			
Ti	47	106.205		
V	51	110.094		
Cr	52	104.882		
Cr	53			
Mn	55	112.401		
Co	59	118.124		
Ni	60	104.355		
Cu	65	104.741		
Zn	66	105.936		
Ge	72		109.004	
As	75	104.876		
Se	82	109.798		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 16:23:06
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.479		
Ag	107	99.972		
Cd	111	98.481		
Cd	114			
In	115		116.255	
Sn	118	99.577		
Sb	123	97.186		
Ba	135	93.802		
Ce	140			
Tb	159			
Ho	165			
Tl	203	102.852		
Tl	205			
Pb	206			
Pb	207			
Pb	208	104.543		
U	238	107.427		
Bi	209		112.161	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Al	27	
QC Std 6	V	51	
QC Std 6	Mn	55	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 16:23:06
Page 3

Approved: July 04, 2016

QC Std 6	Co	59

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 16:24:01
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100067.7	1.3				ug/L	74699	Standard
	Be	9	36.7	7.9	**0.0387**	0.002	5.2	ug/L	8	Standard
	Al	27	10153.6	30.4	**0.0381**	0.015	39.5	ug/L	1493	Standard
	Sc	45	67434.8	2.7				ug/L	61029	Standard
	Ti	47	38.3	21.7	**-0.0212**	0.020	92.7	ug/L	35	Standard
	V	51	1502.7	1.2	**-0.0231**	0.006	25.9	ug/L	1435	Standard
	Cr	52	7451.2	2.4	**-0.1523**	0.012	7.6	ug/L	7635	Standard
	Cr	53	503.3	11.6	**0.1429**	0.063	44.4	ug/L	363	Standard
	Mn	55	3693.5	26.8	**0.3137**	0.134	42.8	ug/L	747	Standard
	Co	59	358.3	10.7	**0.0099**	0.006	62.4	ug/L	234	Standard
	Ni	60	209.7	4.9	**-0.0268**	0.006	22.7	ug/L	215	Standard
	Cu	65	198.7	7.2	**-0.0110**	0.009	80.7	ug/L	104	Standard
	Zn	66	532.3	5.7	**-0.0316**	0.034	106.3	ug/L	363	Standard
>	**Ge**	72	571298.3	2.3				ug/L	495142	Standard
	As	75	-12.4	42.9	**0.0499**	0.004	8.8	ug/L	-50	Standard
	Se	82	23.5	9.9	**0.0476**	0.021	43.8	ug/L	14	Standard
	Se-1	77	58.7	7.7	**0.0749**	0.069	91.8	ug/L	42	Standard
>	**Ga**	71	75.0	26.7				mg/L	42	Standard
	Rb	85	36.7	55.1				ug/L	17	Standard
	Y	89	500557.3	2.5				ug/L	417098	Standard
>	**Rh**	103	30.0	44.1				ug/L	13	Standard
	Mo	98	301.0	26.9	**0.0924**	0.026	27.9	ug/L	11	Standard
	Ag	107	165.0	21.0	**0.0091**	0.004	49.2	ug/L	68	Standard
	Cd	111	19.8	28.3	**0.0003**	0.002	699.6	mg/L	2	Standard
	Cd	114	46.5	85.8	**0.0058**	0.007	116.4	ug/L	12	Standard
>	**In**	115	675355.4	0.8				ug/L	587441	Standard
	Sn	118	876.7	18.6	**0.0359**	0.025	69.7	ug/L	497	Standard
	Sb	123	599.5	22.8	**0.1076**	0.026	24.3	ug/L	78	Standard
	Ba	135	42.7	23.9	**0.0021**	0.004	201.6	ug/L	24	Standard
	Ce	140	140.0	34.4				ug/L	25	Standard
>	**Tb**	159	1044716.9	1.2				ug/L	939701	Standard
	Ho	165	20.0	50.0				ug/L	10	Standard
	Tl	203	103.0	21.9	**0.0094**	0.003	27.0	ug/L	4	Standard
	Tl	205	70.0	58.5	**0.0089**	0.005	57.1	ug/L	2	Standard
	Pb	206	338.7	4.4	**-0.0148**	0.002	16.0	ug/L	240	Standard
	Pb	207	277.7	9.1	**-0.0179**	0.005	28.1	ug/L	198	Standard
	Pb	208	1213.0	9.6	**-0.0165**	0.006	36.1	ug/L	883	Standard
	U	238	62.0	33.9	**0.0036**	0.003	76.1	ug/L	3	Standard
>	**Bi**	209	588682.5	1.2				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 16:26:17
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.0278**	29.484	173.2	mg/L	0	Standard
Mg	24	56.7	20.4	**0.0014**	0.015	1120.4	mg/L	37	Standard
K	39	36.7	28.4	**-0.0407**	0.021	51.2	mg/L	38	Standard
Ca	43	26.7	28.6	**-1.0029**	0.415	41.4	mg/L	10	Standard
Fe	54	187.4	12.0	**0.0055**	0.017	304.1	mg/L	179	Standard
Fe	57	261.7	6.1	**-0.2681**	0.071	26.5	mg/L	335	Standard
Sc-1	45	67434.8	2.7				mg/L	61029	Standard
Cl	35	26219.3	0.7				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	2243.5	3.0				ug/L	1547	Standard
P	31	19277.3	3.9				ug/L	15286	Standard
S	34	2016.8	11.3				ug/L	1572	Standard
Sr	88	70.0	35.7				ug/L	73	Standard
C	12	120.0	25.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	49.4	87.7				mg/L	6	Standard
Ho-1	165	20.0	50.0				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	2125.2	22.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.381	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 16:26:17
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.966
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		112.347
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 16:26:17
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606124701SDL WG574688-02
Sample Date/Time: Wednesday, June 29, 2016 16:27:13
Number of Replicates: 3
Autosampler Position: 215
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	105717.5	0.5				ug/L	74699	Standard
	Be	9	553.3	1.4	**0.4517**	0.004	0.8	ug/L	8	Standard
	Al	27	5662006.5	2.5	**27.2120**	0.787	2.9	ug/L	1493	Standard
	Sc	45	72131.5	3.8				ug/L	61029	Standard
	Ti	47	604.0	7.1	**1.2783**	0.109	8.5	ug/L	35	Standard
	V	51	3941.0	1.4	**0.2585**	0.002	0.9	ug/L	1435	Standard
	Cr	52	12715.3	2.3	**0.4691**	0.023	4.9	ug/L	7635	Standard
	Cr	53	1340.1	9.1	**0.9313**	0.103	11.1	ug/L	363	Standard
	Mn	55	2452895.1	0.8	**306.4368**	3.099	1.0	ug/L	747	Standard
	Co	59	67057.1	0.2	**8.7998**	0.112	1.3	ug/L	234	Standard
	Ni	60	46732.3	0.4	**21.5970**	0.233	1.1	ug/L	215	Standard
	Cu	65	904.0	6.3	**0.3188**	0.021	6.7	ug/L	104	Standard
	Zn	66	62660.6	1.1	**49.4618**	0.122	0.2	ug/L	363	Standard
>	**Ge**	72	569030.6	1.4				ug/L	495142	Standard
	As	75	59.1	19.2	**0.1055**	0.009	8.4	ug/L	-50	Standard
	Se	82	34.4	3.8	**0.1350**	0.014	10.1	ug/L	14	Standard
	Se-1	77	82.7	11.5	**0.3621**	0.106	29.4	ug/L	42	Standard
>	**Ga**	71	288.3	19.2				mg/L	42	Standard
	Rb	85	7325.1	4.9				ug/L	17	Standard
	Y	89	616528.9	0.8				ug/L	417098	Standard
>	**Rh**	103	13.3	78.1				ug/L	13	Standard
	Mo	98	112.4	8.7	**0.0298**	0.003	9.5	ug/L	11	Standard
	Ag	107	122.3	20.4	**0.0029**	0.003	108.6	ug/L	68	Standard
	Cd	111	198.8	14.6	**0.0723**	0.012	16.2	mg/L	2	Standard
	Cd	114	516.3	5.9	**0.0829**	0.006	6.9	ug/L	12	Standard
>	**In**	115	696026.2	0.9				ug/L	587441	Standard
	Sn	118	760.0	4.7	**0.0151**	0.006	37.2	ug/L	497	Standard
	Sb	123	350.2	21.1	**0.0592**	0.014	23.4	ug/L	78	Standard
	Ba	135	775.0	6.3	**0.3035**	0.019	6.3	ug/L	24	Standard
	Ce	140	119180.2	0.3				ug/L	25	Standard
>	**Tb**	159	1103880.2	1.1				ug/L	939701	Standard
	Ho	165	9261.2	2.5				ug/L	10	Standard
	Tl	203	155.7	9.5	**0.0145**	0.001	9.9	ug/L	4	Standard
	Tl	205	156.7	24.4	**0.0191**	0.005	23.7	ug/L	2	Standard
	Pb	206	668.3	5.9	**0.0368**	0.005	14.8	ug/L	240	Standard
	Pb	207	530.0	3.8	**0.0264**	0.004	13.6	ug/L	198	Standard
	Pb	208	2281.7	2.6	**0.0309**	0.002	6.9	ug/L	883	Standard
	U	238	1160.7	2.8	**0.1411**	0.004	2.6	ug/L	3	Standard
>	**Bi**	209	607087.1	0.9				ug/L	523984	Standard

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:29:30
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.5008**	28.572	173.2	mg/L	0	Standard
Mg	24	3728.8	3.6	**4.0090**	0.135	3.4	mg/L	37	Standard
K	39	85.0	42.4	**0.0464**	0.073	156.7	mg/L	38	Standard
Ca	43	183.3	11.4	**7.2814**	1.162	16.0	mg/L	10	Standard
Fe	54	8713.3	2.8	**6.1146**	0.148	2.4	mg/L	179	Standard
Fe	57	2506.9	1.7	**6.1424**	0.321	5.2	mg/L	335	Standard
> **Sc-1**	45	72131.5	3.8				mg/L	61029	Standard
Cl	35	26208.6	1.1				ug/L	20195	Standard
Kr	83	2.7	21.7				ug/L	4	Standard
Br	81	3227.0	4.8				ug/L	1547	Standard
P	31	20363.7	0.5				ug/L	15286	Standard
S	34	1863.4	4.5				ug/L	1572	Standard
Sr	88	105.0	21.8				ug/L	73	Standard
C	12	56.7	36.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	14855.4	4.1				mg/L	6	Standard
Ho-1	165	9261.2	2.5				mg/L	10	Standard
Er	166	8048.8	4.7				mg/L	13	Standard
I	127	5796.1	1.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		141.525	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.923	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:29:30
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	118.484
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	115.860
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606124701SDL WG574688-02
Report Date/Time: Wednesday, June 29, 2016 16:29:30
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126504
Sample Date/Time: Wednesday, June 29, 2016 16:30:25
Number of Replicates: 3
Autosampler Position: 216
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101552.6	1.9				ug/L	74699	Standard
	Be	9	25.0	72.1	**0.0287**	0.016	54.3	ug/L	8	Standard
	Al	27	16250197.0	1.8	**81.3578**	2.887	3.5	ug/L	1493	Standard
	Sc	45	67044.7	1.9				ug/L	61029	Standard
	Ti	47	394.3	10.9	**0.8437**	0.094	11.2	ug/L	35	Standard
	V	51	39093.2	1.3	**4.5488**	0.169	3.7	ug/L	1435	Standard
	Cr	52	19068.3	2.3	**1.3348**	0.101	7.6	ug/L	7635	Standard
	Cr	53	5326.0	20.8	**4.9620**	1.165	23.5	ug/L	363	Standard
	Mn	55	6239.0	2.5	**0.6731**	0.027	4.0	ug/L	747	Standard
	Co	59	851.0	2.5	**0.0808**	0.004	4.5	ug/L	234	Standard
	Ni	60	1951.8	2.1	**0.8314**	0.029	3.5	ug/L	215	Standard
	Cu	65	1321.1	4.3	**0.5464**	0.014	2.5	ug/L	104	Standard
	Zn	66	5985.8	0.6	**4.5673**	0.118	2.6	ug/L	363	Standard
>	**Ge**	72	540512.2	2.5				ug/L	495142	Standard
	As	75	1446.2	5.5	**1.2419**	0.041	3.3	ug/L	-50	Standard
	Se	82	841.6	2.6	**6.8893**	0.246	3.6	ug/L	14	Standard
	Se-1	77	609.0	15.5	**7.0254**	1.243	17.7	ug/L	42	Standard
>	**Ga**	71	105.0	12.6				mg/L	42	Standard
	Rb	85	1541.7	7.2				ug/L	17	Standard
	Y	89	489017.5	2.4				ug/L	417098	Standard
>	**Rh**	103	70.0	18.9				ug/L	13	Standard
	Mo	98	501.8	4.3	**0.1636**	0.005	2.9	ug/L	11	Standard
	Ag	107	97.0	7.4	**0.0005**	0.001	200.8	ug/L	68	Standard
	Cd	111	26.4	23.0	**0.0035**	0.003	79.3	mg/L	2	Standard
	Cd	114	74.1	18.4	**0.0109**	0.002	21.9	ug/L	12	Standard
>	**In**	115	652860.7	1.6				ug/L	587441	Standard
	Sn	118	840.0	9.7	**0.0347**	0.012	35.4	ug/L	497	Standard
	Sb	123	289.8	10.9	**0.0517**	0.007	12.7	ug/L	78	Standard
	Ba	135	17710.3	0.6	**7.7715**	0.161	2.1	ug/L	24	Standard
	Ce	140	1525.1	0.6				ug/L	25	Standard
>	**Tb**	159	1031518.0	1.9				ug/L	939701	Standard
	Ho	165	43.3	24.0				ug/L	10	Standard
	Tl	203	163.3	10.6	**0.0174**	0.002	9.6	ug/L	4	Standard
	Tl	205	133.3	16.9	**0.0182**	0.003	18.7	ug/L	2	Standard
	Pb	206	441.3	2.3	**0.0087**	0.001	8.9	ug/L	240	Standard
	Pb	207	362.0	4.8	**0.0037**	0.003	70.2	ug/L	198	Standard
	Pb	208	1549.4	2.6	**0.0058**	0.001	11.8	ug/L	883	Standard
	U	238	23538.3	0.9	**3.3015**	0.086	2.6	ug/L	3	Standard
>	**Bi**	209	542188.4	2.1				ug/L	523984	Standard

Sample ID: L1606126504
Report Date/Time: Wednesday, June 29, 2016 16:32:42
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**35.7447**	61.903	173.2	mg/L	0	Standard
Mg	24	119069.4	0.9	**139.8814**	3.117	2.2	mg/L	37	Standard
K	39	143.3	11.2	**0.1739**	0.028	15.9	mg/L	38	Standard
Ca	43	463.3	1.6	**24.1242**	0.657	2.7	mg/L	10	Standard
Fe	54	197.6	4.4	**0.0142**	0.006	45.1	mg/L	179	Standard
Fe	57	435.0	13.1	**0.2738**	0.196	71.7	mg/L	335	Standard
> **Sc-1**	45	67044.7	1.9				mg/L	61029	Standard
Cl	35	31470.7	2.7				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	73323.9	4.5				ug/L	1547	Standard
P	31	23608.4	2.5				ug/L	15286	Standard
S	34	1800.1	6.6				ug/L	1572	Standard
Sr	88	293.3	6.9				ug/L	73	Standard
C	12	193.3	41.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	86.5	19.9				mg/L	6	Standard
Ho-1	165	43.3	24.0				mg/L	10	Standard
Er	166	73.3	34.3				mg/L	13	Standard
I	127	54762.3	5.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		135.949	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		109.163	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606126504
Report Date/Time: Wednesday, June 29, 2016 16:32:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.136
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.474
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126505
Sample Date/Time: Wednesday, June 29, 2016 16:33:36
Number of Replicates: 3
Autosampler Position: 217
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	99450.4	1.4				ug/L	74699	Standard
	Be	9	58.3	30.1	**0.0575**	0.016	27.1	ug/L	8	Standard
	Al	27	32118811.5	1.4	**164.1461**	0.108	0.1	ug/L	1493	Standard
	Sc	45	70749.8	1.3				ug/L	61029	Standard
	Ti	47	642.0	2.2	**1.4252**	0.030	2.1	ug/L	35	Standard
	V	51	8228.7	11.8	**0.7915**	0.119	15.1	ug/L	1435	Standard
	Cr	52	11308.2	2.8	**0.3584**	0.044	12.4	ug/L	7635	Standard
	Cr	53	3712.1	10.4	**3.3091**	0.391	11.8	ug/L	363	Standard
	Mn	55	3368829.0	1.1	**438.1329**	6.772	1.5	ug/L	747	Standard
	Co	59	3400.7	28.1	**0.4294**	0.132	30.8	ug/L	234	Standard
	Ni	60	4538.7	3.1	**2.0720**	0.077	3.7	ug/L	215	Standard
	Cu	65	1589.1	4.3	**0.6697**	0.036	5.3	ug/L	104	Standard
	Zn	66	6182.6	5.7	**4.6729**	0.311	6.7	ug/L	363	Standard
>	**Ge**	72	546665.5	0.5				ug/L	495142	Standard
	As	75	6156.5	0.5	**5.0379**	0.006	0.1	ug/L	-50	Standard
	Se	82	164.0	1.8	**1.2151**	0.026	2.2	ug/L	14	Standard
	Se-1	77	258.7	9.7	**2.5862**	0.322	12.5	ug/L	42	Standard
>	**Ga**	71	188.3	13.1				mg/L	42	Standard
	Rb	85	1378.4	10.6				ug/L	17	Standard
	Y	89	484009.6	2.1				ug/L	417098	Standard
>	**Rh**	103	118.3	6.5				ug/L	13	Standard
	Mo	98	286.3	31.1	**0.0925**	0.030	32.6	ug/L	11	Standard
	Ag	107	183.3	43.5	**0.0128**	0.011	86.7	ug/L	68	Standard
	Cd	111	248.2	17.8	**0.1006**	0.018	18.1	mg/L	2	Standard
	Cd	114	590.9	17.2	**0.1032**	0.017	16.3	ug/L	12	Standard
>	**In**	115	642069.8	1.9				ug/L	587441	Standard
	Sn	118	836.7	5.6	**0.0365**	0.010	26.8	ug/L	497	Standard
	Sb	123	288.3	29.8	**0.0524**	0.017	32.5	ug/L	78	Standard
	Ba	135	131296.8	1.0	**58.6819**	0.595	1.0	ug/L	24	Standard
	Ce	140	6513.1	10.5				ug/L	25	Standard
>	**Tb**	159	1012113.9	1.4				ug/L	939701	Standard
	Ho	165	216.7	7.1				ug/L	10	Standard
	Tl	203	185.3	61.4	**0.0202**	0.013	66.8	ug/L	4	Standard
	Tl	205	276.7	88.2	**0.0379**	0.033	87.8	ug/L	2	Standard
	Pb	206	891.4	31.8	**0.0916**	0.052	56.6	ug/L	240	Standard
	Pb	207	748.7	30.8	**0.0839**	0.048	56.6	ug/L	198	Standard
	Pb	208	3050.1	30.7	**0.0846**	0.049	57.6	ug/L	883	Standard
	U	238	21997.6	2.6	**3.1135**	0.091	2.9	ug/L	3	Standard
>	**Bi**	209	537104.2	0.4				ug/L	523984	Standard

Sample ID: L1606126505
Report Date/Time: Wednesday, June 29, 2016 16:35:53
Page 1

Approved: July 04, 2016

Na	23	13.3	57.3	**132.8617**	77.017	58.0	mg/L	0	Standard
Mg	24	104917.1	1.8	**116.7540**	0.597	0.5	mg/L	37	Standard
K	39	116.7	8.9	**0.1082**	0.018	16.4	mg/L	38	Standard
Ca	43	916.7	6.0	**47.4532**	3.610	7.6	mg/L	10	Standard
Fe	54	3013.9	2.4	**2.0666**	0.083	4.0	mg/L	179	Standard
Fe	57	1381.7	4.1	**2.9770**	0.162	5.4	mg/L	335	Standard
> **Sc-1**	45	70749.8	1.3				mg/L	61029	Standard
Cl	35	30215.4	1.0				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	35768.9	2.6				ug/L	1547	Standard
P	31	25429.7	2.5				ug/L	15286	Standard
S	34	2313.5	0.3				ug/L	1572	Standard
Sr	88	551.7	8.2				ug/L	73	Standard
C	12	433.3	25.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	3	Standard
Dy	164	279.9	12.3				mg/L	6	Standard
Ho-1	165	216.7	7.1				mg/L	10	Standard
Er	166	280.0	18.9				mg/L	13	Standard
I	127	794211.5	6.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		133.135	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.406	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126505
Report Date/Time: Wednesday, June 29, 2016 16:35:53
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	109.299
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	102.504
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606126505
Report Date/Time: Wednesday, June 29, 2016 16:35:53
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126506
Sample Date/Time: Wednesday, June 29, 2016 16:36:48
Number of Replicates: 3
Autosampler Position: 218
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100252.3	1.1				ug/L	74699	Standard
	Be	9	26.7	39.0	**0.0303**	0.009	29.8	ug/L	8	Standard
	Al	27	69444.0	4.9	**0.3387**	0.017	5.1	ug/L	1493	Standard
	Sc	45	71205.3	1.8				ug/L	61029	Standard
	Ti	47	113.0	20.9	**0.1414**	0.055	38.5	ug/L	35	Standard
	V	51	2901.2	2.9	**0.1267**	0.013	10.0	ug/L	1435	Standard
	Cr	52	23756.9	1.1	**1.6651**	0.003	0.2	ug/L	7635	Standard
	Cr	53	2710.2	4.1	**2.1308**	0.127	5.9	ug/L	363	Standard
	Mn	55	7460.8	10.3	**0.7511**	0.098	13.0	ug/L	747	Standard
	Co	59	456.0	8.6	**0.0207**	0.006	26.8	ug/L	234	Standard
	Ni	60	2497.2	2.2	**0.9949**	0.023	2.3	ug/L	215	Standard
	Cu	65	984.0	1.2	**0.3398**	0.007	2.0	ug/L	104	Standard
	Zn	66	8303.3	0.9	**5.9218**	0.058	1.0	ug/L	363	Standard
>	**Ge**	72	590298.4	1.1				ug/L	495142	Standard
	As	75	7.9	237.8	**0.0655**	0.014	21.6	ug/L	-50	Standard
	Se	82	27.7	14.2	**0.0742**	0.032	43.2	ug/L	14	Standard
	Se-1	77	72.3	15.2	**0.2091**	0.134	63.9	ug/L	42	Standard
>	**Ga**	71	91.7	20.7				mg/L	42	Standard
	Rb	85	308.3	4.1				ug/L	17	Standard
	Y	89	498995.2	1.7				ug/L	417098	Standard
>	**Rh**	103	21.7	48.0				ug/L	13	Standard
	Mo	98	716.7	9.5	**0.2259**	0.023	10.1	ug/L	11	Standard
	Ag	107	155.3	16.4	**0.0076**	0.003	44.6	ug/L	68	Standard
	Cd	111	44.4	12.9	**0.0104**	0.002	21.8	mg/L	2	Standard
	Cd	114	117.5	18.7	**0.0177**	0.004	21.5	ug/L	12	Standard
>	**In**	115	682064.0	0.6				ug/L	587441	Standard
	Sn	118	4689.1	2.4	**0.5966**	0.013	2.3	ug/L	497	Standard
	Sb	123	228.7	11.3	**0.0380**	0.005	12.5	ug/L	78	Standard
	Ba	135	1034.0	4.7	**0.4191**	0.020	4.8	ug/L	24	Standard
	Ce	140	486.7	14.8				ug/L	25	Standard
>	**Tb**	159	1056147.0	0.9				ug/L	939701	Standard
	Ho	165	8.3	69.3				ug/L	10	Standard
	Tl	203	43.7	26.0	**0.0029**	0.001	43.8	ug/L	4	Standard
	Tl	205	50.0	26.5	**0.0063**	0.002	26.5	ug/L	2	Standard
	Pb	206	665.7	3.1	**0.0376**	0.003	7.4	ug/L	240	Standard
	Pb	207	542.7	5.2	**0.0299**	0.006	20.6	ug/L	198	Standard
	Pb	208	2227.4	4.5	**0.0297**	0.006	20.4	ug/L	883	Standard
	U	238	38.7	20.9	**0.0005**	0.001	221.1	ug/L	3	Standard
>	**Bi**	209	599996.8	1.5				ug/L	523984	Standard

Sample ID: L1606126506
Report Date/Time: Wednesday, June 29, 2016 16:39:04
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	216.7	5.8	**0.1745**	0.010	5.5	mg/L	37	Standard
K	39	46.7	49.5	**-0.0257**	0.044	170.8	mg/L	38	Standard
Ca	43	30.0	16.7	**-0.8971**	0.290	32.3	mg/L	10	Standard
Fe	54	200.7	16.8	**0.0076**	0.025	326.7	mg/L	179	Standard
Fe	57	248.3	11.1	**-0.3516**	0.067	19.2	mg/L	335	Standard
> **Sc-1**	45	71205.3	1.8				mg/L	61029	Standard
Cl	35	31458.6	1.1				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	3110.3	10.8				ug/L	1547	Standard
P	31	21037.9	1.6				ug/L	15286	Standard
S	34	2263.5	3.0				ug/L	1572	Standard
Sr	88	58.3	27.6				ug/L	73	Standard
C	12	113.3	13.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	38.9	27.0				mg/L	6	Standard
Ho-1	165	8.3	69.3				mg/L	10	Standard
Er	166	23.3	65.5				mg/L	13	Standard
I	127	20210.2	61.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		134.209	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		119.218	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126506
Report Date/Time: Wednesday, June 29, 2016 16:39:04
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	116.108
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	114.507
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126601
Sample Date/Time: Wednesday, June 29, 2016 16:39:59
Number of Replicates: 3
Autosampler Position: 219
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101321.2	3.1				ug/L	74699	Standard
	Be	9	18.3	56.8	**0.0228**	0.008	36.0	ug/L	8	Standard
	Al	27	44431569.1	0.7	**222.9926**	5.415	2.4	ug/L	1493	Standard
	Sc	45	71310.8	2.2				ug/L	61029	Standard
	Ti	47	275.3	6.3	**0.5272**	0.047	8.9	ug/L	35	Standard
	V	51	27907.2	1.9	**3.0387**	0.030	1.0	ug/L	1435	Standard
	Cr	52	13580.1	0.8	**0.5803**	0.015	2.6	ug/L	7635	Standard
	Cr	53	2821.9	2.8	**2.3430**	0.081	3.4	ug/L	363	Standard
	Mn	55	153037.1	1.2	**19.0926**	0.078	0.4	ug/L	747	Standard
	Co	59	1012.0	1.5	**0.0969**	0.001	1.5	ug/L	234	Standard
	Ni	60	4370.3	2.4	**1.9194**	0.048	2.5	ug/L	215	Standard
	Cu	65	751.7	5.3	**0.2499**	0.020	8.0	ug/L	104	Standard
	Zn	66	3518.1	1.4	**2.3650**	0.050	2.1	ug/L	363	Standard
>	**Ge**	72	565652.8	1.1				ug/L	495142	Standard
	As	75	1269.0	4.3	**1.0512**	0.035	3.4	ug/L	-50	Standard
	Se	82	382.5	3.3	**2.9130**	0.125	4.3	ug/L	14	Standard
	Se-1	77	193.3	14.8	**1.6975**	0.369	21.7	ug/L	42	Standard
>	**Ga**	71	66.7	18.9				mg/L	42	Standard
	Rb	85	2323.5	7.7				ug/L	17	Standard
	Y	89	480038.2	1.2				ug/L	417098	Standard
>	**Rh**	103	150.0	3.3				ug/L	13	Standard
	Mo	98	638.7	4.4	**0.2084**	0.010	5.0	ug/L	11	Standard
	Ag	107	92.0	14.1	**-0.0003**	0.002	569.4	ug/L	68	Standard
	Cd	111	195.2	3.6	**0.0755**	0.003	3.5	mg/L	2	Standard
	Cd	114	469.8	19.5	**0.0797**	0.016	19.7	ug/L	12	Standard
>	**In**	115	657497.5	0.6				ug/L	587441	Standard
	Sn	118	753.4	8.5	**0.0205**	0.010	48.9	ug/L	497	Standard
	Sb	123	222.8	13.6	**0.0384**	0.006	15.3	ug/L	78	Standard
	Ba	135	100054.6	0.7	**43.6617**	0.503	1.2	ug/L	24	Standard
	Ce	140	173.3	6.0				ug/L	25	Standard
>	**Tb**	159	1045560.4	1.5				ug/L	939701	Standard
	Ho	165	40.0	25.0				ug/L	10	Standard
	Tl	203	129.3	1.6	**0.0127**	0.000	3.7	ug/L	4	Standard
	Tl	205	133.3	20.7	**0.0174**	0.003	20.0	ug/L	2	Standard
	Pb	206	412.3	3.7	**0.0001**	0.002	2456.7	ug/L	240	Standard
	Pb	207	343.7	8.9	**-0.0032**	0.005	165.7	ug/L	198	Standard
	Pb	208	1432.7	1.3	**-0.0035**	0.002	56.3	ug/L	883	Standard
	U	238	43636.7	0.7	**5.8505**	0.120	2.1	ug/L	3	Standard
>	**Bi**	209	567479.2	1.8				ug/L	523984	Standard

Sample ID: L1606126601
Report Date/Time: Wednesday, June 29, 2016 16:42:16
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**48.6438**	48.507	99.7	mg/L	0	Standard
Mg	24	35550.0	1.9	**39.2140**	0.699	1.8	mg/L	37	Standard
K	39	165.0	24.1	**0.1975**	0.071	36.0	mg/L	38	Standard
Ca	43	596.7	5.4	**29.7553**	2.186	7.3	mg/L	10	Standard
Fe	54	190.1	10.6	**-0.0004**	0.014	3403.7	mg/L	179	Standard
Fe	57	555.0	7.7	**0.5391**	0.093	17.2	mg/L	335	Standard
> **Sc-1**	45	71310.8	2.2				mg/L	61029	Standard
Cl	35	33374.1	0.9				ug/L	20195	Standard
Kr	83	2.0	132.3				ug/L	4	Standard
Br	81	77209.6	1.2				ug/L	1547	Standard
P	31	24828.7	1.5				ug/L	15286	Standard
S	34	2263.5	4.7				ug/L	1572	Standard
Sr	88	591.7	14.4				ug/L	73	Standard
C	12	253.3	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	40.0	25.0				mg/L	3	Standard
Dy	164	40.1	59.8				mg/L	6	Standard
Ho-1	165	40.0	25.0				mg/L	10	Standard
Er	166	66.7	37.7				mg/L	13	Standard
I	127	2058609.0	7.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		135.640	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.240	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126601
Report Date/Time: Wednesday, June 29, 2016 16:42:16
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.926
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.301
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606126901
Sample Date/Time: Wednesday, June 29, 2016 16:43:10
Number of Replicates: 3
Autosampler Position: 220
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	97179.0	1.2				ug/L	74699	Standard
	Be	9	23.3	44.6	**0.0281**	0.009	33.4	ug/L	8	Standard
	Al	27	25054309.4	1.7	**131.0332**	1.604	1.2	ug/L	1493	Standard
	Sc	45	69363.5	2.7				ug/L	61029	Standard
	Ti	47	323.3	2.8	**0.6992**	0.021	2.9	ug/L	35	Standard
	V	51	12024.0	2.6	**1.3128**	0.037	2.9	ug/L	1435	Standard
	Cr	52	11840.6	1.6	**0.4900**	0.025	5.0	ug/L	7635	Standard
	Cr	53	2901.9	3.9	**2.6451**	0.119	4.5	ug/L	363	Standard
	Mn	55	490032.6	0.8	**66.5395**	0.816	1.2	ug/L	747	Standard
	Co	59	931.7	2.1	**0.0964**	0.003	3.3	ug/L	234	Standard
	Ni	60	2217.5	3.2	**0.9981**	0.030	3.0	ug/L	215	Standard
	Cu	65	599.0	3.0	**0.2013**	0.011	5.4	ug/L	104	Standard
	Zn	66	4806.1	1.2	**3.7141**	0.039	1.0	ug/L	363	Standard
>	**Ge**	72	522601.4	0.6				ug/L	495142	Standard
	As	75	2361.3	3.3	**2.0571**	0.073	3.5	ug/L	-50	Standard
	Se	82	593.9	4.2	**4.9896**	0.230	4.6	ug/L	14	Standard
	Se-1	77	179.7	2.8	**1.7086**	0.076	4.5	ug/L	42	Standard
>	**Ga**	71	105.0	9.5				mg/L	42	Standard
	Rb	85	2805.3	2.5				ug/L	17	Standard
	Y	89	474068.5	3.1				ug/L	417098	Standard
>	**Rh**	103	81.7	12.7				ug/L	13	Standard
	Mo	98	305.2	1.9	**0.0985**	0.004	4.4	ug/L	11	Standard
	Ag	107	99.7	12.7	**0.0010**	0.002	153.4	ug/L	68	Standard
	Cd	111	19.2	16.0	**0.0004**	0.001	278.0	mg/L	2	Standard
	Cd	114	49.5	42.7	**0.0067**	0.004	59.0	ug/L	12	Standard
>	**In**	115	645532.6	2.3				ug/L	587441	Standard
	Sn	118	888.4	9.9	**0.0435**	0.010	24.0	ug/L	497	Standard
	Sb	123	240.8	4.4	**0.0427**	0.002	4.3	ug/L	78	Standard
	Ba	135	155622.1	0.5	**69.2020**	1.756	2.5	ug/L	24	Standard
	Ce	140	928.4	7.9				ug/L	25	Standard
>	**Tb**	159	1004174.2	1.0				ug/L	939701	Standard
	Ho	165	23.3	32.7				ug/L	10	Standard
	Tl	203	115.7	10.5	**0.0121**	0.001	11.5	ug/L	4	Standard
	Tl	205	128.3	22.2	**0.0179**	0.004	22.0	ug/L	2	Standard
	Pb	206	395.0	0.9	**0.0016**	0.001	57.1	ug/L	240	Standard
	Pb	207	323.3	6.0	**-0.0029**	0.004	138.0	ug/L	198	Standard
	Pb	208	1415.7	2.0	**0.0002**	0.001	432.7	ug/L	883	Standard
	U	238	7265.7	2.0	**1.0348**	0.023	2.2	ug/L	3	Standard
>	**Bi**	209	532226.3	0.6				ug/L	523984	Standard

Sample ID: L1606126901
Report Date/Time: Wednesday, June 29, 2016 16:45:27
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.4177**	30.160	173.2	mg/L	0	Standard
Mg	24	119202.1	1.4	**135.3768**	3.695	2.7	mg/L	37	Standard
K	39	220.0	8.2	**0.3143**	0.042	13.4	mg/L	38	Standard
Ca	43	398.3	4.4	**19.6130**	0.698	3.6	mg/L	10	Standard
Fe	54	252.1	18.1	**0.0493**	0.029	59.6	mg/L	179	Standard
Fe	57	418.3	9.7	**0.1785**	0.138	77.4	mg/L	335	Standard
> **Sc-1**	45	69363.5	2.7				mg/L	61029	Standard
Cl	35	32454.7	1.8				ug/L	20195	Standard
Kr	83	1.0	173.2				ug/L	4	Standard
Br	81	137582.4	3.7				ug/L	1547	Standard
P	31	24828.7	1.5				ug/L	15286	Standard
S	34	2306.8	3.7				ug/L	1572	Standard
Sr	88	291.7	6.0				ug/L	73	Standard
C	12	323.3	19.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	15.2	73.1				mg/L	6	Standard
Ho-1	165	23.3	32.7				mg/L	10	Standard
Er	166	30.0	33.3				mg/L	13	Standard
I	127	310633.7	2.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		130.094	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		105.546	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606126901
Report Date/Time: Wednesday, June 29, 2016 16:45:27
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.889
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		101.573
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127201
Sample Date/Time: Wednesday, June 29, 2016 16:46:22
Number of Replicates: 3
Autosampler Position: 221
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94337.9	2.3				ug/L	74699	Standard
	Be	9	23.3	53.9	**0.0288**	0.012	40.8	ug/L	8	Standard
	Al	27	2978340.1	2.2	**16.0395**	0.522	3.3	ug/L	1493	Standard
	Sc	45	67137.0	4.7				ug/L	61029	Standard
	Ti	47	236.7	18.3	**0.4520**	0.099	21.9	ug/L	35	Standard
	V	51	33734.3	0.8	**3.8201**	0.030	0.8	ug/L	1435	Standard
	Cr	52	13200.4	3.0	**0.5779**	0.062	10.7	ug/L	7635	Standard
	Cr	53	1396.7	5.5	**1.0289**	0.077	7.5	ug/L	363	Standard
	Mn	55	5810.0	55.0	**0.6034**	0.416	68.9	ug/L	747	Standard
	Co	59	597.0	23.2	**0.0441**	0.019	43.4	ug/L	234	Standard
	Ni	60	839.7	12.2	**0.2798**	0.055	19.7	ug/L	215	Standard
	Cu	65	499.7	4.1	**0.1378**	0.013	9.3	ug/L	104	Standard
	Zn	66	1881.8	10.3	**1.0958**	0.181	16.5	ug/L	363	Standard
>	**Ge**	72	550779.1	1.5				ug/L	495142	Standard
	As	75	823.5	3.7	**0.7205**	0.021	2.9	ug/L	-50	Standard
	Se	82	534.3	3.8	**4.2403**	0.211	5.0	ug/L	14	Standard
	Se-1	77	394.7	5.2	**4.2338**	0.183	4.3	ug/L	42	Standard
>	**Ga**	71	76.7	13.6				mg/L	42	Standard
	Rb	85	1120.0	6.4				ug/L	17	Standard
	Y	89	478775.5	2.0				ug/L	417098	Standard
>	**Rh**	103	33.3	52.7				ug/L	13	Standard
	Mo	98	1510.8	4.0	**0.5152**	0.028	5.4	ug/L	11	Standard
	Ag	107	155.0	42.6	**0.0091**	0.010	106.9	ug/L	68	Standard
	Cd	111	66.6	64.6	**0.0216**	0.019	89.7	mg/L	2	Standard
	Cd	114	188.5	50.4	**0.0319**	0.018	55.0	ug/L	12	Standard
>	**In**	115	639482.4	2.0				ug/L	587441	Standard
	Sn	118	738.4	13.9	**0.0215**	0.017	79.5	ug/L	497	Standard
	Sb	123	301.8	22.4	**0.0554**	0.014	26.0	ug/L	78	Standard
	Ba	135	10964.6	1.7	**4.9054**	0.035	0.7	ug/L	24	Standard
	Ce	140	728.4	65.5				ug/L	25	Standard
>	**Tb**	159	995838.4	2.1				ug/L	939701	Standard
	Ho	165	55.0	78.7				ug/L	10	Standard
	Tl	203	165.3	64.1	**0.0175**	0.012	69.7	ug/L	4	Standard
	Tl	205	88.3	34.1	**0.0121**	0.004	34.3	ug/L	2	Standard
	Pb	206	371.3	17.5	**-0.0043**	0.012	275.3	ug/L	240	Standard
	Pb	207	297.3	13.9	**-0.0098**	0.008	82.9	ug/L	198	Standard
	Pb	208	1253.7	10.6	**-0.0098**	0.007	68.1	ug/L	883	Standard
	U	238	7715.3	1.4	**1.0741**	0.027	2.5	ug/L	3	Standard
>	**Bi**	209	544657.1	1.1				ug/L	523984	Standard

Sample ID: L1606127201
Report Date/Time: Wednesday, June 29, 2016 16:48:39
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**87.3928**	30.523	34.9	mg/L	0	Standard
Mg	24	56818.0	2.0	**66.6625**	1.829	2.7	mg/L	37	Standard
K	39	118.3	28.8	**0.1220**	0.059	48.6	mg/L	38	Standard
Ca	43	173.3	14.2	**7.4463**	1.537	20.6	mg/L	10	Standard
Fe	54	200.1	8.9	**0.0164**	0.019	116.3	mg/L	179	Standard
Fe	57	341.7	12.8	**-0.0156**	0.156	997.8	mg/L	335	Standard
> **Sc-1**	45	67137.0	4.7				mg/L	61029	Standard
Cl	35	31870.8	2.6				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	13232.4	3.4				ug/L	1547	Standard
P	31	22137.8	2.1				ug/L	15286	Standard
S	34	2160.2	8.0				ug/L	1572	Standard
Sr	88	146.7	24.2				ug/L	73	Standard
C	12	170.0	10.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	64.3	100.4				mg/L	6	Standard
Ho-1	165	55.0	78.7				mg/L	10	Standard
Er	166	50.0	60.0				mg/L	13	Standard
I	127	62898.8	4.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		126.291	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.237	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127201
Report Date/Time: Wednesday, June 29, 2016 16:48:39
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.859
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.945
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127202
Sample Date/Time: Wednesday, June 29, 2016 16:49:33
Number of Replicates: 3
Autosampler Position: 222
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98788.0	4.1				ug/L	74699	Standard
	Be	9	13.3	57.3	**0.0192**	0.007	36.1	ug/L	8	Standard
	Al	27	43121483.9	2.2	**221.9923**	5.444	2.5	ug/L	1493	Standard
	Sc	45	66930.9	3.1				ug/L	61029	Standard
	Ti	47	1039.7	26.4	**2.4141**	0.661	27.4	ug/L	35	Standard
	V	51	49957.2	0.9	**5.8910**	0.067	1.1	ug/L	1435	Standard
	Cr	52	18344.1	1.8	**1.2538**	0.038	3.0	ug/L	7635	Standard
	Cr	53	24960.6	4.5	**24.5202**	1.080	4.4	ug/L	363	Standard
	Mn	55	17873.9	4.1	**2.2143**	0.103	4.7	ug/L	747	Standard
	Co	59	1405.4	6.7	**0.1586**	0.013	8.4	ug/L	234	Standard
	Ni	60	3876.2	1.9	**1.7810**	0.032	1.8	ug/L	215	Standard
	Cu	65	1407.4	2.7	**0.5921**	0.020	3.4	ug/L	104	Standard
	Zn	66	4623.4	0.6	**3.4401**	0.014	0.4	ug/L	363	Standard
>	**Ge**	72	538103.3	0.3				ug/L	495142	Standard
	As	75	2665.5	0.4	**2.2493**	0.008	0.3	ug/L	-50	Standard
	Se	82	859.9	5.4	**7.0717**	0.401	5.7	ug/L	14	Standard
	Se-1	77	1900.1	5.6	**23.3152**	1.302	5.6	ug/L	42	Standard
>	**Ga**	71	205.0	14.8				mg/L	42	Standard
	Rb	85	12111.5	1.9				ug/L	17	Standard
	Y	89	447538.5	3.5				ug/L	417098	Standard
>	**Rh**	103	186.7	5.6				ug/L	13	Standard
	Mo	98	4087.3	1.0	**1.4942**	0.026	1.7	ug/L	11	Standard
	Ag	107	258.7	10.6	**0.0259**	0.004	14.5	ug/L	68	Standard
	Cd	111	70.2	15.3	**0.0250**	0.005	21.6	mg/L	2	Standard
	Cd	114	224.5	9.2	**0.0407**	0.004	8.7	ug/L	12	Standard
>	**In**	115	600654.0	1.1				ug/L	587441	Standard
	Sn	118	765.0	14.6	**0.0335**	0.020	60.1	ug/L	497	Standard
	Sb	123	166.1	4.9	**0.0306**	0.002	5.1	ug/L	78	Standard
	Ba	135	52090.1	1.0	**24.8744**	0.038	0.2	ug/L	24	Standard
	Ce	140	4474.0	3.6				ug/L	25	Standard
>	**Tb**	159	974974.5	0.9				ug/L	939701	Standard
	Ho	165	80.0	27.2				ug/L	10	Standard
	Tl	203	82.0	10.0	**0.0082**	0.001	13.1	ug/L	4	Standard
	Tl	205	63.3	43.5	**0.0090**	0.004	42.2	ug/L	2	Standard
	Pb	206	1123.7	3.1	**0.1383**	0.006	4.7	ug/L	240	Standard
	Pb	207	914.0	2.3	**0.1218**	0.004	3.1	ug/L	198	Standard
	Pb	208	3779.9	2.5	**0.1267**	0.006	5.0	ug/L	883	Standard
	U	238	23648.8	1.1	**3.4170**	0.057	1.7	ug/L	3	Standard
>	**Bi**	209	526202.0	0.7				ug/L	523984	Standard

Sample ID: L1606127202
Report Date/Time: Wednesday, June 29, 2016 16:51:50
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**53.6294**	53.947	100.6	mg/L	0	Standard
Mg	24	113104.0	1.5	**133.1465**	4.642	3.5	mg/L	37	Standard
K	39	478.3	11.2	**0.8485**	0.082	9.7	mg/L	38	Standard
Ca	43	695.0	6.9	**37.5535**	3.460	9.2	mg/L	10	Standard
Fe	54	250.5	4.6	**0.0553**	0.004	6.4	mg/L	179	Standard
Fe	57	631.7	3.6	**0.8833**	0.015	1.7	mg/L	335	Standard
> **Sc-1**	45	66930.9	3.1				mg/L	61029	Standard
Cl	35	32927.1	2.7				ug/L	20195	Standard
Kr	83	2.7	57.3				ug/L	4	Standard
Br	81	41968.1	3.2				ug/L	1547	Standard
P	31	23413.1	2.8				ug/L	15286	Standard
S	34	2041.8	4.5				ug/L	1572	Standard
Sr	88	750.0	13.9				ug/L	73	Standard
C	12	126.7	19.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	103.6	24.2				mg/L	6	Standard
Ho-1	165	80.0	27.2				mg/L	10	Standard
Er	166	63.3	32.9				mg/L	13	Standard
I	127	36619.2	2.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		132.248	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.676	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127202
Report Date/Time: Wednesday, June 29, 2016 16:51:50
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.249
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.423
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127901
Sample Date/Time: Wednesday, June 29, 2016 16:52:45
Number of Replicates: 3
Autosampler Position: 223
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94669.8	0.8				ug/L	74699	Standard
	Be	9	153.3	14.7	**0.1451**	0.021	14.4	ug/L	8	Standard
	Al	27	36828943.8	1.6	**197.7220**	2.539	1.3	ug/L	1493	Standard
	Sc	45	72494.8	2.9				ug/L	61029	Standard
	Ti	47	9823.2	1.5	**22.6070**	0.583	2.6	ug/L	35	Standard
	V	51	40160.9	1.2	**4.4642**	0.016	0.4	ug/L	1435	Standard
	Cr	52	82051.3	1.3	**8.6794**	0.190	2.2	ug/L	7635	Standard
	Cr	53	10588.7	4.0	**9.7108**	0.425	4.4	ug/L	363	Standard
	Mn	55	3156665.4	1.9	**397.2480**	11.605	2.9	ug/L	747	Standard
	Co	59	13202.1	1.7	**1.7150**	0.048	2.8	ug/L	234	Standard
	Ni	60	14584.3	1.0	**6.7029**	0.116	1.7	ug/L	215	Standard
	Cu	65	6796.9	2.1	**3.0972**	0.099	3.2	ug/L	104	Standard
	Zn	66	29601.9	0.7	**23.2962**	0.407	1.7	ug/L	363	Standard
>	**Ge**	72	565029.1	1.1				ug/L	495142	Standard
	As	75	1169.4	2.0	**0.9746**	0.024	2.5	ug/L	-50	Standard
	Se	82	54.4	2.8	**0.2964**	0.014	4.6	ug/L	14	Standard
	Se-1	77	95.3	8.8	**0.5209**	0.088	16.8	ug/L	42	Standard
>	**Ga**	71	2745.3	10.7				mg/L	42	Standard
	Rb	85	40779.8	2.2				ug/L	17	Standard
	Y	89	501434.2	0.6				ug/L	417098	Standard
>	**Rh**	103	116.7	40.7				ug/L	13	Standard
	Mo	98	6510.4	0.7	**2.1843**	0.030	1.4	ug/L	11	Standard
	Ag	107	146.0	1.2	**0.0072**	0.000	3.6	ug/L	68	Standard
	Cd	111	54.1	19.4	**0.0153**	0.004	28.2	mg/L	2	Standard
	Cd	114	169.8	20.2	**0.0276**	0.006	22.4	ug/L	12	Standard
>	**In**	115	655252.5	1.2				ug/L	587441	Standard
	Sn	118	1435.1	4.4	**0.1255**	0.007	5.6	ug/L	497	Standard
	Sb	123	419.8	7.1	**0.0764**	0.005	6.8	ug/L	78	Standard
	Ba	135	167537.6	1.4	**73.3690**	0.619	0.8	ug/L	24	Standard
	Ce	140	144147.5	0.8				ug/L	25	Standard
>	**Tb**	159	1041877.9	1.3				ug/L	939701	Standard
	Ho	165	2731.9	2.9				ug/L	10	Standard
	Tl	203	237.3	5.5	**0.0244**	0.001	5.3	ug/L	4	Standard
	Tl	205	228.3	8.3	**0.0292**	0.002	8.4	ug/L	2	Standard
	Pb	206	7357.1	0.9	**1.1818**	0.018	1.5	ug/L	240	Standard
	Pb	207	5920.5	2.7	**1.0648**	0.026	2.5	ug/L	198	Standard
	Pb	208	25273.6	1.1	**1.1521**	0.016	1.4	ug/L	883	Standard
	U	238	16614.4	0.8	**2.1915**	0.022	1.0	ug/L	3	Standard
>	**Bi**	209	575969.2	0.6				ug/L	523984	Standard

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:55:02
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**32.3968**	28.087	86.7	mg/L	0	Standard
Mg	24	3568.8	4.7	**3.8119**	0.091	2.4	mg/L	37	Standard
K	39	285.0	24.6	**0.4167**	0.132	31.6	mg/L	38	Standard
Ca	43	908.4	6.1	**45.7783**	2.516	5.5	mg/L	10	Standard
Fe	54	3605.6	1.3	**2.4364**	0.072	2.9	mg/L	179	Standard
Fe	57	1446.7	0.5	**3.0674**	0.109	3.5	mg/L	335	Standard
Sc-1	45	72494.8	2.9				mg/L	61029	Standard
Cl	35	31293.6	1.2				ug/L	20195	Standard
Kr	83	2.3	49.5				ug/L	4	Standard
Br	81	8015.5	4.3				ug/L	1547	Standard
P	31	26808.7	3.1				ug/L	15286	Standard
S	34	2353.5	0.9				ug/L	1572	Standard
Sr	88	506.7	5.4				ug/L	73	Standard
C	12	183.3	20.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	3	Standard
Dy	164	4522.3	4.2				mg/L	6	Standard
Ho-1	165	2731.9	2.9				mg/L	10	Standard
Er	166	2333.5	8.8				mg/L	13	Standard
I	127	38263.4	7.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		126.735	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.114	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:55:02
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.544
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.921
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:55:02
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127901
Sample Date/Time: Wednesday, June 29, 2016 16:55:56
Number of Replicates: 3
Autosampler Position: 224
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	94757.1	0.6				ug/L	74699	Standard
	Be	9	16.7	34.6	**0.0226**	0.005	23.2	ug/L	8	Standard
	Al	27	809583.1	0.4	**4.3294**	0.010	0.2	ug/L	1493	Standard
	Sc	45	70063.3	2.2				ug/L	61029	Standard
	Ti	47	239.7	2.8	**0.4418**	0.016	3.7	ug/L	35	Standard
	V	51	2635.2	3.3	**0.1083**	0.010	9.6	ug/L	1435	Standard
	Cr	52	10594.0	0.1	**0.2214**	0.004	1.7	ug/L	7635	Standard
	Cr	53	1061.7	10.6	**0.6708**	0.107	16.0	ug/L	363	Standard
	Mn	55	69843.0	1.9	**8.5894**	0.152	1.8	ug/L	747	Standard
	Co	59	472.7	1.3	**0.0251**	0.001	3.3	ug/L	234	Standard
	Ni	60	497.3	0.6	**0.1075**	0.002	1.4	ug/L	215	Standard
	Cu	65	314.7	2.4	**0.0436**	0.003	7.6	ug/L	104	Standard
	Zn	66	3346.7	3.2	**2.2142**	0.089	4.0	ug/L	363	Standard
>	**Ge**	72	568460.2	0.3				ug/L	495142	Standard
	As	75	-20.7	45.1	**0.0435**	0.007	16.6	ug/L	-50	Standard
	Se	82	18.3	17.6	**0.0072**	0.025	350.3	ug/L	14	Standard
	Se-1	77	58.0	6.2	**0.0694**	0.043	62.2	ug/L	42	Standard
>	**Ga**	71	123.3	16.9				mg/L	42	Standard
	Rb	85	875.0	6.6				ug/L	17	Standard
	Y	89	490331.7	1.2				ug/L	417098	Standard
>	**Rh**	103	16.7	45.8				ug/L	13	Standard
	Mo	98	175.2	3.1	**0.0513**	0.001	2.5	ug/L	11	Standard
	Ag	107	187.0	5.8	**0.0120**	0.001	10.3	ug/L	68	Standard
	Cd	111	8.4	36.2	**-0.0044**	0.001	27.6	mg/L	2	Standard
	Cd	114	20.0	63.9	**0.0013**	0.002	167.4	ug/L	12	Standard
>	**In**	115	676596.2	0.9				ug/L	587441	Standard
	Sn	118	621.7	13.0	**-0.0023**	0.012	549.8	ug/L	497	Standard
	Sb	123	148.0	17.5	**0.0233**	0.005	19.5	ug/L	78	Standard
	Ba	135	3597.4	2.3	**1.5099**	0.025	1.6	ug/L	24	Standard
	Ce	140	3133.7	5.7				ug/L	25	Standard
>	**Tb**	159	1053037.0	0.7				ug/L	939701	Standard
	Ho	165	73.3	37.6				ug/L	10	Standard
	Tl	203	38.7	9.8	**0.0024**	0.000	14.0	ug/L	4	Standard
	Tl	205	30.0	57.7	**0.0039**	0.002	55.8	ug/L	2	Standard
	Pb	206	500.3	0.4	**0.0113**	0.001	11.7	ug/L	240	Standard
	Pb	207	394.7	10.5	**0.0033**	0.008	256.3	ug/L	198	Standard
	Pb	208	1728.4	3.9	**0.0071**	0.003	37.0	ug/L	883	Standard
	U	238	362.3	7.5	**0.0420**	0.004	8.9	ug/L	3	Standard
>	**Bi**	209	594905.3	1.7				ug/L	523984	Standard

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:58:13
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.1214**	29.647	173.2	mg/L	0	Standard
Mg	24	156.7	18.1	**0.1107**	0.028	25.7	mg/L	37	Standard
K	39	30.0	60.1	**-0.0567**	0.034	59.1	mg/L	38	Standard
Ca	43	53.3	5.4	**0.4139**	0.213	51.5	mg/L	10	Standard
Fe	54	254.1	20.9	**0.0493**	0.038	78.1	mg/L	179	Standard
Fe	57	286.7	7.0	**-0.2249**	0.071	31.4	mg/L	335	Standard
> **Sc-1**	45	70063.3	2.2				mg/L	61029	Standard
Cl	35	33357.3	0.7				ug/L	20195	Standard
Kr	83	2.3	89.2				ug/L	4	Standard
Br	81	2703.6	11.1				ug/L	1547	Standard
P	31	21573.7	6.4				ug/L	15286	Standard
S	34	2518.5	1.8				ug/L	1572	Standard
Sr	88	90.0	20.0				ug/L	73	Standard
C	12	76.7	7.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	206.2	21.0				mg/L	6	Standard
Ho-1	165	73.3	37.6				mg/L	10	Standard
Er	166	80.0	12.5				mg/L	13	Standard
I	127	6318.0	20.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		126.852	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.807	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127901
Report Date/Time: Wednesday, June 29, 2016 16:58:13
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		115.177
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.535
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 16:59:09
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92097.4	0.2				ug/L	74699	Standard
	Be	9	59332.7	2.3	**54.6708**	1.340	2.5	ug/L	8	Standard
	Al	27	8920018.4	1.0	**49.2175**	0.589	1.2	ug/L	1493	Standard
	Sc	45	68374.0	2.7				ug/L	61029	Standard
	Ti	47	44746.6	0.7	**103.2626**	1.432	1.4	ug/L	35	Standard
	V	51	467593.6	0.9	**54.0107**	0.396	0.7	ug/L	1435	Standard
	Cr	52	444978.1	1.4	**51.5506**	1.111	2.2	ug/L	7635	Standard
	Cr	53	54491.2	2.0	**51.2973**	1.411	2.7	ug/L	363	Standard
	Mn	55	420710.6	2.8	**52.7641**	1.799	3.4	ug/L	747	Standard
	Co	59	425350.8	2.7	**56.3649**	1.831	3.2	ug/L	234	Standard
	Ni	60	109363.3	1.2	**51.0215**	0.971	1.9	ug/L	215	Standard
	Cu	65	108714.3	1.3	**51.0413**	1.018	2.0	ug/L	104	Standard
	Zn	66	63872.0	1.2	**50.7455**	1.020	2.0	ug/L	363	Standard
>	**Ge**	72	565538.7	0.9				ug/L	495142	Standard
	As	75	65590.6	1.0	**51.3298**	0.623	1.2	ug/L	-50	Standard
	Se	82	6676.5	1.4	**53.1245**	0.865	1.6	ug/L	14	Standard
	Se-1	77	4570.7	2.8	**54.1793**	1.989	3.7	ug/L	42	Standard
>	**Ga**	71	90.0	25.5				mg/L	42	Standard
	Rb	85	980.0	1.8				ug/L	17	Standard
	Y	89	481101.3	1.3				ug/L	417098	Standard
>	**Rh**	103	36.7	20.8				ug/L	13	Standard
	Mo	98	317636.1	2.3	**106.0178**	2.920	2.8	ug/L	11	Standard
	Ag	107	377887.3	1.0	**51.5689**	0.593	1.2	ug/L	68	Standard
	Cd	111	119799.3	1.9	**50.9674**	1.189	2.3	mg/L	2	Standard
	Cd	114	291345.5	1.8	**50.5384**	1.150	2.3	ug/L	12	Standard
>	**In**	115	660345.6	0.5				ug/L	587441	Standard
	Sn	118	333887.3	1.0	**50.7546**	0.756	1.5	ug/L	497	Standard
	Sb	123	263770.3	1.5	**50.2579**	0.987	2.0	ug/L	78	Standard
	Ba	135	110751.0	1.4	**48.1238**	0.926	1.9	ug/L	24	Standard
	Ce	140	100.0	13.2				ug/L	25	Standard
>	**Tb**	159	1027328.3	1.0				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	469667.3	0.3	**51.2282**	0.784	1.5	ug/L	4	Standard
	Tl	205	432915.4	1.7	**54.4135**	1.621	3.0	ug/L	2	Standard
	Pb	206	304012.7	1.3	**51.1045**	1.311	2.6	ug/L	240	Standard
	Pb	207	269196.3	1.3	**50.9328**	1.301	2.6	ug/L	198	Standard
	Pb	208	1105976.1	1.5	**52.9652**	1.506	2.8	ug/L	883	Standard
	U	238	408284.4	0.8	**53.3305**	1.093	2.0	ug/L	3	Standard
>	**Bi**	209	582837.4	1.3				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:01:26
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4664.1	3.2	**5.3139**	0.299	5.6	mg/L	37	Standard
K	39	2066.8	2.7	**3.9658**	0.176	4.4	mg/L	38	Standard
Ca	43	83.3	3.5	**2.1752**	0.111	5.1	mg/L	10	Standard
Fe	54	7259.4	3.7	**5.3547**	0.102	1.9	mg/L	179	Standard
Fe	57	2016.8	2.5	**5.0496**	0.304	6.0	mg/L	335	Standard
Sc-1	45	68374.0	2.7				mg/L	61029	Standard
Cl	35	31290.3	2.5				ug/L	20195	Standard
Kr	83	2.3	65.5				ug/L	4	Standard
Br	81	2203.5	6.4				ug/L	1547	Standard
P	31	22518.4	0.7				ug/L	15286	Standard
S	34	2795.3	6.4				ug/L	1572	Standard
Sr	88	58.3	13.1				ug/L	73	Standard
C	12	70.0	37.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	25.1	45.0				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	33.3	34.6				mg/L	13	Standard
I	127	4404.0	17.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	109.342		
Al	27	98.435		
Sc	45			
Ti	47	103.263		
V	51	108.021		
Cr	52	103.101		
Cr	53			
Mn	55	105.528		
Co	59	112.730		
Ni	60	102.043		
Cu	65	102.083		
Zn	66	101.491		
Ge	72		114.217	
As	75	102.660		
Se	82	106.249		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:01:26
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	106.018		
Ag	107	103.138		
Cd	111	101.935		
Cd	114			
In	115		112.411	
Sn	118	101.509		
Sb	123	100.516		
Ba	135	96.248		
Ce	140			
Tb	159			
Ho	165			
Tl	203	102.456		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.930		
U	238	106.661		
Bi	209		111.232	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:01:26
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 17:02:20
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92622.3	1.7				ug/L	74699	Standard
	Be	9	38.3	64.3	**0.0426**	0.023	52.8	ug/L	8	Standard
	Al	27	23370.4	151.2	**0.1159**	0.196	168.9	ug/L	1493	Standard
	Sc	45	68951.6	1.3				ug/L	61029	Standard
	Ti	47	40.3	21.1	**-0.0139**	0.019	138.6	ug/L	35	Standard
	V	51	1665.7	22.9	**0.0015**	0.044	2840.4	ug/L	1435	Standard
	Cr	52	7712.6	2.1	**-0.0923**	0.036	39.5	ug/L	7635	Standard
	Cr	53	583.3	10.9	**0.2339**	0.049	21.0	ug/L	363	Standard
	Mn	55	4624.3	138.3	**0.4433**	0.816	184.2	ug/L	747	Standard
	Co	59	398.3	65.9	**0.0165**	0.035	212.9	ug/L	234	Standard
	Ni	60	246.3	49.7	**-0.0062**	0.058	932.9	ug/L	215	Standard
	Cu	65	154.7	27.9	**-0.0294**	0.020	68.5	ug/L	104	Standard
	Zn	66	470.3	19.7	**-0.0693**	0.075	107.9	ug/L	363	Standard
>	**Ge**	72	553538.3	2.4				ug/L	495142	Standard
	As	75	-38.0	63.4	**0.0295**	0.019	62.8	ug/L	-50	Standard
	Se	82	20.6	12.7	**0.0307**	0.026	83.7	ug/L	14	Standard
	Se-1	77	63.0	12.0	**0.1490**	0.085	57.3	ug/L	42	Standard
>	**Ga**	71	71.7	14.5				mg/L	42	Standard
	Rb	85	25.0	0.0				ug/L	17	Standard
	Y	89	484828.2	0.4				ug/L	417098	Standard
>	**Rh**	103	11.7	65.5				ug/L	13	Standard
	Mo	98	326.7	26.1	**0.0989**	0.027	26.9	ug/L	11	Standard
	Ag	107	215.0	76.0	**0.0152**	0.021	139.6	ug/L	68	Standard
	Cd	111	16.1	49.4	**-0.0013**	0.003	242.1	mg/L	2	Standard
	Cd	114	45.0	78.6	**0.0054**	0.006	109.2	ug/L	12	Standard
>	**In**	115	687277.7	0.8				ug/L	587441	Standard
	Sn	118	931.7	18.6	**0.0418**	0.027	63.5	ug/L	497	Standard
	Sb	123	489.1	40.1	**0.0853**	0.036	41.8	ug/L	78	Standard
	Ba	135	52.0	47.0	**0.0057**	0.010	177.6	ug/L	24	Standard
	Ce	140	81.7	78.0				ug/L	25	Standard
>	**Tb**	159	1059717.9	1.6				ug/L	939701	Standard
	Ho	165	6.7	86.6				ug/L	10	Standard
	Tl	203	97.3	80.9	**0.0086**	0.008	97.3	ug/L	4	Standard
	Tl	205	126.7	77.5	**0.0158**	0.012	76.4	ug/L	2	Standard
	Pb	206	434.7	47.9	**0.0006**	0.034	5998.1	ug/L	240	Standard
	Pb	207	327.7	41.4	**-0.0091**	0.025	275.6	ug/L	198	Standard
	Pb	208	1259.4	26.8	**-0.0148**	0.016	107.4	ug/L	883	Standard
	U	238	94.7	104.8	**0.0077**	0.013	165.1	ug/L	3	Standard
>	**Bi**	209	593862.7	2.1				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:04:37
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	88.3	93.4	**0.0352**	0.093	262.9	mg/L	37	Standard
K	39	41.7	30.2	**-0.0325**	0.025	77.1	mg/L	38	Standard
Ca	43	21.7	58.1	**-1.3088**	0.708	54.1	mg/L	10	Standard
Fe	54	169.3	10.4	**-0.0113**	0.012	105.6	mg/L	179	Standard
Fe	57	255.0	7.1	**-0.3067**	0.061	20.0	mg/L	335	Standard
> **Sc-1**	45	68951.6	1.3				mg/L	61029	Standard
Cl	35	30862.1	1.0				ug/L	20195	Standard
Kr	83	3.7	31.5				ug/L	4	Standard
Br	81	2130.2	6.5				ug/L	1547	Standard
P	31	22533.4	1.3				ug/L	15286	Standard
S	34	2576.9	2.2				ug/L	1572	Standard
Sr	88	61.7	20.4				ug/L	73	Standard
C	12	70.0	62.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	15.1	95.9				mg/L	6	Standard
Ho-1	165	6.7	86.6				mg/L	10	Standard
Er	166	33.3	69.3				mg/L	13	Standard
I	127	3348.7	10.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.794	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:04:37
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.995
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.336
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Mn	55	

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:04:37
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127902 WG574001-03
Sample Date/Time: Wednesday, June 29, 2016 17:05:33
Number of Replicates: 3
Autosampler Position: 225
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	97844.9	1.2				ug/L	74699	Standard
	Be	9	628.3	5.4	**0.5525**	0.030	5.3	ug/L	8	Standard
	Al	27	67592650.7	0.8	**351.1766**	6.777	1.9	ug/L	1493	Standard
	Sc	45	76196.2	3.2				ug/L	61029	Standard
	Ti	47	35031.5	1.8	**85.4603**	1.454	1.7	ug/L	35	Standard
	V	51	127107.8	0.4	**15.3856**	0.087	0.6	ug/L	1435	Standard
	Cr	52	144594.0	0.9	**17.0395**	0.185	1.1	ug/L	7635	Standard
	Cr	53	17708.7	3.4	**17.4084**	0.576	3.3	ug/L	363	Standard
	Mn	55	9638149.6	0.5	**1281.4698**	6.113	0.5	ug/L	747	Standard
	Co	59	44221.4	1.1	**6.1626**	0.079	1.3	ug/L	234	Standard
	Ni	60	29440.2	0.7	**14.4336**	0.102	0.7	ug/L	215	Standard
	Cu	65	18715.2	1.1	**9.2058**	0.118	1.3	ug/L	104	Standard
	Zn	66	49911.7	1.0	**41.8483**	0.443	1.1	ug/L	363	Standard
>	**Ge**	72	534821.8	0.2				ug/L	495142	Standard
	As	75	5363.3	0.7	**4.4926**	0.039	0.9	ug/L	-50	Standard
	Se	82	81.7	4.2	**0.5509**	0.030	5.5	ug/L	14	Standard
	Se-1	77	90.0	5.1	**0.5184**	0.056	10.8	ug/L	42	Standard
>	**Ga**	71	10246.8	1.6				mg/L	42	Standard
	Rb	85	128176.6	1.2				ug/L	17	Standard
	Y	89	611826.2	1.2				ug/L	417098	Standard
>	**Rh**	103	203.3	5.1				ug/L	13	Standard
	Mo	98	8459.2	1.1	**2.8837**	0.063	2.2	ug/L	11	Standard
	Ag	107	335.7	32.7	**0.0339**	0.015	43.7	ug/L	68	Standard
	Cd	111	167.4	13.2	**0.0649**	0.009	13.6	mg/L	2	Standard
	Cd	114	471.4	29.4	**0.0814**	0.024	29.1	ug/L	12	Standard
>	**In**	115	645352.1	1.1				ug/L	587441	Standard
	Sn	118	2433.5	1.0	**0.2845**	0.006	2.0	ug/L	497	Standard
	Sb	123	747.2	5.1	**0.1415**	0.006	4.4	ug/L	78	Standard
	Ba	135	186708.8	1.1	**83.0208**	0.490	0.6	ug/L	24	Standard
	Ce	140	555105.9	0.4				ug/L	25	Standard
>	**Tb**	159	1017797.9	1.6				ug/L	939701	Standard
	Ho	165	10682.1	1.8				ug/L	10	Standard
	Tl	203	801.4	12.7	**0.0906**	0.011	12.4	ug/L	4	Standard
	Tl	205	726.7	5.1	**0.0967**	0.005	5.6	ug/L	2	Standard
	Pb	206	28687.5	0.5	**5.0315**	0.023	0.5	ug/L	240	Standard
	Pb	207	23046.8	1.0	**4.5438**	0.005	0.1	ug/L	198	Standard
	Pb	208	100519.8	0.6	**5.0199**	0.074	1.5	ug/L	883	Standard
	U	238	32058.9	0.6	**4.4211**	0.050	1.1	ug/L	3	Standard
>	**Bi**	209	551473.5	0.9				ug/L	523984	Standard

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:07:50
Page 1

Approved: July 04, 2016

Na	23	16.7	17.3	**153.4719**	22.118	14.4	mg/L	0	Standard
Mg	24	8050.5	4.4	**8.2571**	0.176	2.1	mg/L	37	Standard
K	39	990.0	1.3	**1.6393**	0.041	2.5	mg/L	38	Standard
Ca	43	1383.4	8.1	**67.3924**	3.470	5.1	mg/L	10	Standard
Fe	54	12019.5	0.9	**8.0301**	0.303	3.8	mg/L	179	Standard
Fe	57	3888.8	1.5	**9.5221**	0.470	4.9	mg/L	335	Standard
Sc-1	45	76196.2	3.2				mg/L	61029	Standard
Cl	35	38282.3	2.8				ug/L	20195	Standard
Kr	83	1.7	124.9				ug/L	4	Standard
Br	81	9239.5	4.1				ug/L	1547	Standard
P	31	33428.5	3.8				ug/L	15286	Standard
S	34	2046.8	10.4				ug/L	1572	Standard
Sr	88	885.0	8.4				ug/L	73	Standard
C	12	223.3	22.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	3	Standard
Dy	164	18454.9	2.8				mg/L	6	Standard
Ho-1	165	10682.1	1.8				mg/L	10	Standard
Er	166	9956.6	4.7				mg/L	13	Standard
I	127	65381.5	9.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		130.986	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		108.014	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:07:50
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.858
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.246
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:07:50
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127902 WG574001-03
Sample Date/Time: Wednesday, June 29, 2016 17:08:45
Number of Replicates: 3
Autosampler Position: 226
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	97333.3	1.1				ug/L	74699	Standard
	Be	9	23.3	32.7	**0.0279**	0.007	23.3	ug/L	8	Standard
	Al	27	1471688.9	2.0	**7.6717**	0.099	1.3	ug/L	1493	Standard
	Sc	45	68584.9	0.8				ug/L	61029	Standard
	Ti	47	736.0	3.1	**1.5714**	0.076	4.9	ug/L	35	Standard
	V	51	4360.2	2.0	**0.3040**	0.018	5.9	ug/L	1435	Standard
	Cr	52	10891.2	1.4	**0.2475**	0.011	4.5	ug/L	7635	Standard
	Cr	53	901.7	3.1	**0.5139**	0.018	3.4	ug/L	363	Standard
	Mn	55	209342.7	1.0	**25.8667**	0.657	2.5	ug/L	747	Standard
	Co	59	1184.0	8.5	**0.1180**	0.016	13.3	ug/L	234	Standard
	Ni	60	823.4	2.4	**0.2566**	0.015	5.7	ug/L	215	Standard
	Cu	65	604.0	5.8	**0.1771**	0.017	9.7	ug/L	104	Standard
	Zn	66	3536.4	2.1	**2.3459**	0.038	1.6	ug/L	363	Standard
>	**Ge**	72	572433.6	1.6				ug/L	495142	Standard
	As	75	56.5	19.8	**0.1032**	0.008	8.2	ug/L	-50	Standard
	Se	82	22.7	8.9	**0.0409**	0.015	36.5	ug/L	14	Standard
	Se-1	77	56.3	3.7	**0.0451**	0.032	71.6	ug/L	42	Standard
>	**Ga**	71	286.7	20.8				mg/L	42	Standard
	Rb	85	2816.9	8.1				ug/L	17	Standard
	Y	89	483993.7	0.1				ug/L	417098	Standard
>	**Rh**	103	11.7	65.5				ug/L	13	Standard
	Mo	98	255.8	9.6	**0.0784**	0.008	10.6	ug/L	11	Standard
	Ag	107	215.7	39.7	**0.0161**	0.012	72.2	ug/L	68	Standard
	Cd	111	33.2	125.7	**0.0061**	0.018	290.4	mg/L	2	Standard
	Cd	114	83.8	140.0	**0.0123**	0.020	164.2	ug/L	12	Standard
>	**In**	115	670122.5	0.4				ug/L	587441	Standard
	Sn	118	730.0	9.6	**0.0149**	0.011	73.4	ug/L	497	Standard
	Sb	123	198.5	22.7	**0.0331**	0.009	26.1	ug/L	78	Standard
	Ba	135	4104.6	4.5	**1.7422**	0.086	5.0	ug/L	24	Standard
	Ce	140	11919.7	3.4				ug/L	25	Standard
>	**Tb**	159	1040073.5	0.3				ug/L	939701	Standard
	Ho	165	248.3	25.2				ug/L	10	Standard
	Tl	203	119.3	122.5	**0.0111**	0.016	142.6	ug/L	4	Standard
	Tl	205	45.0	44.4	**0.0058**	0.003	43.8	ug/L	2	Standard
	Pb	206	1027.4	11.1	**0.0985**	0.020	20.7	ug/L	240	Standard
	Pb	207	828.0	1.8	**0.0839**	0.004	4.9	ug/L	198	Standard
	Pb	208	3389.1	2.4	**0.0853**	0.005	6.3	ug/L	883	Standard
	U	238	690.0	3.8	**0.0840**	0.004	4.3	ug/L	3	Standard
>	**Bi**	209	594228.5	1.0				ug/L	523984	Standard

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:11:02
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	215.0	6.2	**0.1819**	0.016	8.8	mg/L	37	Standard
K	39	53.3	14.3	**-0.0093**	0.014	153.2	mg/L	38	Standard
Ca	43	48.3	15.8	**0.1914**	0.411	214.9	mg/L	10	Standard
Fe	54	443.5	15.8	**0.1966**	0.055	28.1	mg/L	179	Standard
Fe	57	323.3	11.7	**-0.0960**	0.118	122.9	mg/L	335	Standard
> **Sc-1**	45	68584.9	0.8				mg/L	61029	Standard
Cl	35	29711.8	0.5				ug/L	20195	Standard
Kr	83	0.7	173.2				ug/L	4	Standard
Br	81	2166.8	7.4				ug/L	1547	Standard
P	31	19419.1	3.8				ug/L	15286	Standard
S	34	2206.8	3.0				ug/L	1572	Standard
Sr	88	96.7	19.6				ug/L	73	Standard
C	12	46.7	49.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	390.3	4.3				mg/L	6	Standard
Ho-1	165	248.3	25.2				mg/L	10	Standard
Er	166	203.3	36.3				mg/L	13	Standard
I	127	5617.8	25.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		130.301	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.610	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127902 WG574001-03
Report Date/Time: Wednesday, June 29, 2016 17:11:02
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.075
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.406
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127903S WG574001-09
Sample Date/Time: Wednesday, June 29, 2016 17:11:56
Number of Replicates: 3
Autosampler Position: 227
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	103734.0	1.8				ug/L	74699	Standard
	Be	9	62659.4	2.5	**51.2580**	0.858	1.7	ug/L	8	Standard
	Al	27	65559814.6	2.9	**321.3254**	12.360	3.8	ug/L	1493	Standard
	Sc	45	78765.9	1.8				ug/L	61029	Standard
	Ti	47	33652.0	1.4	**81.0287**	2.211	2.7	ug/L	35	Standard
	V	51	549023.7	0.9	**66.2300**	1.368	2.1	ug/L	1435	Standard
	Cr	52	510507.6	0.8	**61.9303**	1.817	2.9	ug/L	7635	Standard
	Cr	53	62490.3	1.3	**61.4602**	1.770	2.9	ug/L	363	Standard
	Mn	55	9311676.0	0.4	**1221.9970**	25.998	2.1	ug/L	747	Standard
	Co	59	420012.9	0.8	**58.0870**	1.492	2.6	ug/L	234	Standard
	Ni	60	121605.8	1.9	**59.2404**	2.250	3.8	ug/L	215	Standard
	Cu	65	110409.9	1.5	**54.1128**	1.822	3.4	ug/L	104	Standard
	Zn	66	101498.2	0.9	**84.4641**	2.491	2.9	ug/L	363	Standard
>	**Ge**	72	542031.1	2.4				ug/L	495142	Standard
	As	75	66455.1	1.4	**54.2648**	0.548	1.0	ug/L	-50	Standard
	Se	82	6239.8	1.8	**51.8120**	1.305	2.5	ug/L	14	Standard
	Se-1	77	4136.2	2.8	**51.1231**	1.591	3.1	ug/L	42	Standard
>	**Ga**	71	10863.9	0.3				mg/L	42	Standard
	Rb	85	119392.1	2.4				ug/L	17	Standard
	Y	89	586391.1	1.2				ug/L	417098	Standard
>	**Rh**	103	175.0	15.1				ug/L	13	Standard
	Mo	98	5740.7	2.3	**1.9719**	0.058	3.0	ug/L	11	Standard
	Ag	107	333249.6	2.0	**46.9378**	1.203	2.6	ug/L	68	Standard
	Cd	111	109090.4	0.8	**47.8987**	0.737	1.5	mg/L	2	Standard
	Cd	114	263561.8	0.5	**47.1881**	0.999	2.1	ug/L	12	Standard
>	**In**	115	639905.2	1.9				ug/L	587441	Standard
	Sn	118	1925.1	5.8	**0.2078**	0.017	8.3	ug/L	497	Standard
	Sb	123	124661.9	1.5	**24.5159**	0.713	2.9	ug/L	78	Standard
	Ba	135	268706.4	1.4	**120.5245**	2.161	1.8	ug/L	24	Standard
	Ce	140	538323.1	0.7				ug/L	25	Standard
>	**Tb**	159	1045169.0	2.0				ug/L	939701	Standard
	Ho	165	11317.6	5.4				ug/L	10	Standard
	Tl	203	438870.4	1.4	**49.1992**	0.166	0.3	ug/L	4	Standard
	Tl	205	398898.5	1.5	**51.5243**	0.281	0.5	ug/L	2	Standard
	Pb	206	314887.1	0.5	**54.4073**	0.494	0.9	ug/L	240	Standard
	Pb	207	267442.5	0.8	**52.0088**	0.664	1.3	ug/L	198	Standard
	Pb	208	1101699.4	1.8	**54.2208**	0.536	1.0	ug/L	883	Standard
	U	238	415059.0	1.7	**55.7208**	0.743	1.3	ug/L	3	Standard
>	**Bi**	209	566992.1	1.1				ug/L	523984	Standard

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:14:13
Page 1

Approved: July 04, 2016

Na	23	8.3	69.3	**74.8950**	52.957	70.7	mg/L	0	Standard
Mg	24	7490.2	5.9	**7.4319**	0.549	7.4	mg/L	37	Standard
K	39	995.0	6.5	**1.5912**	0.134	8.4	mg/L	38	Standard
Ca	43	1293.4	1.8	**60.7979**	1.957	3.2	mg/L	10	Standard
Fe	54	10550.1	1.1	**6.7940**	0.164	2.4	mg/L	179	Standard
Fe	57	3448.7	1.2	**8.0106**	0.227	2.8	mg/L	335	Standard
> **Sc-1**	45	78765.9	1.8				mg/L	61029	Standard
Cl	35	36168.5	1.3				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	7995.5	8.9				ug/L	1547	Standard
P	31	34405.7	2.4				ug/L	15286	Standard
S	34	1880.1	7.8				ug/L	1572	Standard
Sr	88	881.7	7.6				ug/L	73	Standard
C	12	140.0	46.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	3	Standard
Dy	164	18849.9	4.6				mg/L	6	Standard
Ho-1	165	11317.6	5.4				mg/L	10	Standard
Er	166	9863.2	2.6				mg/L	13	Standard
I	127	61177.5	10.8				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		138.870	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		109.470	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:14:13
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	108.931
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.208
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:14:13
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127903S WG574001-09
Sample Date/Time: Wednesday, June 29, 2016 17:15:08
Number of Replicates: 3
Autosampler Position: 228
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100014.2	2.8				ug/L	74699	Standard
	Be	9	1316.7	11.4	**1.1231**	0.100	8.9	ug/L	8	Standard
	Al	27	1420391.6	2.6	**7.2057**	0.041	0.6	ug/L	1493	Standard
	Sc	45	65355.8	3.5				ug/L	61029	Standard
	Ti	47	751.7	2.2	**1.7435**	0.043	2.4	ug/L	35	Standard
	V	51	12553.2	2.5	**1.3568**	0.030	2.2	ug/L	1435	Standard
	Cr	52	18458.6	1.1	**1.3028**	0.035	2.7	ug/L	7635	Standard
	Cr	53	1771.8	3.3	**1.4619**	0.098	6.7	ug/L	363	Standard
	Mn	55	203580.1	1.5	**27.1635**	0.320	1.2	ug/L	747	Standard
	Co	59	9135.1	1.0	**1.2550**	0.021	1.7	ug/L	234	Standard
	Ni	60	2779.3	1.4	**1.2628**	0.031	2.5	ug/L	215	Standard
	Cu	65	2563.2	1.7	**1.1828**	0.027	2.3	ug/L	104	Standard
	Zn	66	4447.0	0.1	**3.3488**	0.082	2.5	ug/L	363	Standard
>	**Ge**	72	530183.0	2.3				ug/L	495142	Standard
	As	75	1256.2	4.8	**1.1077**	0.065	5.9	ug/L	-50	Standard
	Se	82	144.5	11.0	**1.0936**	0.160	14.6	ug/L	14	Standard
	Se-1	77	149.3	5.6	**1.2873**	0.101	7.8	ug/L	42	Standard
>	**Ga**	71	291.7	12.0				mg/L	42	Standard
	Rb	85	2473.5	8.7				ug/L	17	Standard
	Y	89	480063.9	1.1				ug/L	417098	Standard
>	**Rh**	103	18.3	15.7				ug/L	13	Standard
	Mo	98	168.4	5.0	**0.0496**	0.003	5.9	ug/L	11	Standard
	Ag	107	7406.8	1.4	**0.9815**	0.012	1.2	ug/L	68	Standard
	Cd	111	2324.9	2.0	**0.9649**	0.013	1.3	mg/L	2	Standard
	Cd	114	5499.5	2.5	**0.9362**	0.028	2.9	ug/L	12	Standard
>	**In**	115	671407.0	1.0				ug/L	587441	Standard
	Sn	118	800.0	6.0	**0.0251**	0.007	29.1	ug/L	497	Standard
	Sb	123	3310.0	4.5	**0.6160**	0.025	4.0	ug/L	78	Standard
	Ba	135	5738.4	0.8	**2.4371**	0.009	0.4	ug/L	24	Standard
	Ce	140	11352.6	1.4				ug/L	25	Standard
>	**Tb**	159	1030086.3	0.3				ug/L	939701	Standard
	Ho	165	295.0	7.8				ug/L	10	Standard
	Tl	203	9225.5	0.2	**0.9995**	0.013	1.3	ug/L	4	Standard
	Tl	205	8459.0	5.0	**1.0575**	0.041	3.9	ug/L	2	Standard
	Pb	206	6958.9	0.8	**1.0943**	0.025	2.3	ug/L	240	Standard
	Pb	207	5899.8	0.2	**1.0421**	0.014	1.4	ug/L	198	Standard
	Pb	208	25022.5	1.7	**1.1197**	0.025	2.3	ug/L	883	Standard
	U	238	8748.9	2.2	**1.1326**	0.025	2.2	ug/L	3	Standard
>	**Bi**	209	585775.3	1.4				ug/L	523984	Standard

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:17:25
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	185.0	5.4	**0.1579**	0.009	5.9	mg/L	37	Standard
K	39	46.7	32.7	**-0.0170**	0.035	202.9	mg/L	38	Standard
Ca	43	53.3	5.4	**0.6286**	0.271	43.1	mg/L	10	Standard
Fe	54	400.2	11.9	**0.1796**	0.048	26.7	mg/L	179	Standard
Fe	57	335.0	9.8	**-0.0119**	0.080	674.7	mg/L	335	Standard
Sc-1	45	65355.8	3.5				mg/L	61029	Standard
Cl	35	29320.3	1.0				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	2030.1	9.5				ug/L	1547	Standard
P	31	19806.3	2.6				ug/L	15286	Standard
S	34	2145.2	0.2				ug/L	1572	Standard
Sr	88	86.7	23.3				ug/L	73	Standard
C	12	80.0	66.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	379.6	21.3				mg/L	6	Standard
Ho-1	165	295.0	7.8				mg/L	10	Standard
Er	166	216.7	11.6				mg/L	13	Standard
I	127	5157.6	28.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		133.890	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		107.077	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127903S WG574001-09
Report Date/Time: Wednesday, June 29, 2016 17:17:25
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	114.294
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	111.792
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127904SD WG574001-10
Sample Date/Time: Wednesday, June 29, 2016 17:18:19
Number of Replicates: 3
Autosampler Position: 229
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	100168.4	0.8				ug/L	74699	Standard
	Be	9	61913.0	2.3	**52.4502**	1.133	2.2	ug/L	8	Standard
	Al	27	66193036.9	1.4	**335.9039**	6.576	2.0	ug/L	1493	Standard
	Sc	45	75343.6	2.4				ug/L	61029	Standard
	Ti	47	40945.9	1.4	**103.9053**	1.926	1.9	ug/L	35	Standard
	V	51	552663.1	1.5	**70.2550**	1.056	1.5	ug/L	1435	Standard
	Cr	52	495290.7	1.7	**63.3235**	1.419	2.2	ug/L	7635	Standard
	Cr	53	60987.5	1.0	**63.2038**	0.948	1.5	ug/L	363	Standard
	Mn	55	9453870.8	1.1	**1307.2105**	20.907	1.6	ug/L	747	Standard
	Co	59	421300.3	0.1	**61.3867**	0.363	0.6	ug/L	234	Standard
	Ni	60	119269.8	2.5	**61.2106**	1.745	2.9	ug/L	215	Standard
	Cu	65	106997.2	1.7	**55.2474**	1.210	2.2	ug/L	104	Standard
	Zn	66	101243.5	1.4	**88.7847**	1.659	1.9	ug/L	363	Standard
>	**Ge**	72	514295.3	0.5				ug/L	495142	Standard
	As	75	64887.3	0.7	**55.8337**	0.652	1.2	ug/L	-50	Standard
	Se	82	6128.3	0.4	**53.6210**	0.432	0.8	ug/L	14	Standard
	Se-1	77	4025.2	1.7	**52.4403**	1.142	2.2	ug/L	42	Standard
>	**Ga**	71	13474.3	0.3				mg/L	42	Standard
	Rb	85	148814.4	3.2				ug/L	17	Standard
	Y	89	590511.6	1.2				ug/L	417098	Standard
>	**Rh**	103	170.0	19.3				ug/L	13	Standard
	Mo	98	5249.8	1.8	**1.8847**	0.047	2.5	ug/L	11	Standard
	Ag	107	319729.4	0.9	**47.0680**	0.306	0.7	ug/L	68	Standard
	Cd	111	105379.0	1.4	**48.3587**	0.323	0.7	mg/L	2	Standard
	Cd	114	254720.6	1.7	**47.6605**	0.543	1.1	ug/L	12	Standard
>	**In**	115	612131.8	1.0				ug/L	587441	Standard
	Sn	118	2201.8	4.9	**0.2670**	0.017	6.2	ug/L	497	Standard
	Sb	123	135513.9	0.7	**27.8528**	0.396	1.4	ug/L	78	Standard
	Ba	135	276700.5	0.9	**129.7213**	0.066	0.1	ug/L	24	Standard
	Ce	140	594096.9	1.6				ug/L	25	Standard
>	**Tb**	159	982831.9	0.9				ug/L	939701	Standard
	Ho	165	12233.3	1.6				ug/L	10	Standard
	Tl	203	424338.5	0.5	**50.3728**	0.553	1.1	ug/L	4	Standard
	Tl	205	393687.9	2.5	**53.8449**	1.302	2.4	ug/L	2	Standard
	Pb	206	310823.5	0.5	**56.8661**	0.171	0.3	ug/L	240	Standard
	Pb	207	265328.4	0.5	**54.6347**	0.327	0.6	ug/L	198	Standard
	Pb	208	1090751.7	0.6	**56.8486**	0.519	0.9	ug/L	883	Standard
	U	238	409441.4	1.4	**58.2033**	0.887	1.5	ug/L	3	Standard
>	**Bi**	209	535481.7	0.7				ug/L	523984	Standard

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:20:36
Page 1

Approved: July 04, 2016

Na	23	6.7	86.6	**61.5240**	53.288	86.6	mg/L	0	Standard
Mg	24	7358.5	1.6	**7.6312**	0.188	2.5	mg/L	37	Standard
K	39	1028.4	3.5	**1.7267**	0.031	1.8	mg/L	38	Standard
Ca	43	1340.1	2.9	**66.0398**	1.115	1.7	mg/L	10	Standard
Fe	54	11628.0	2.4	**7.8474**	0.058	0.7	mg/L	179	Standard
Fe	57	3898.8	5.2	**9.6559**	0.353	3.7	mg/L	335	Standard
Sc-1	45	75343.6	2.4				mg/L	61029	Standard
Cl	35	34624.2	2.4				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	10260.1	1.8				ug/L	1547	Standard
P	31	37270.8	0.5				ug/L	15286	Standard
S	34	1845.1	3.3				ug/L	1572	Standard
Sr	88	923.4	9.7				ug/L	73	Standard
C	12	203.3	29.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	16.7	91.7				mg/L	3	Standard
Dy	164	20358.1	1.7				mg/L	6	Standard
Ho-1	165	12233.3	1.6				mg/L	10	Standard
Er	166	11334.2	2.0				mg/L	13	Standard
I	127	62766.8	7.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		134.096	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		103.868	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:20:36
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		104.203
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.194
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Ti 47 Upper, S, EEE	Ti	47	

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:20:36
Page 3

Approved: July 04, 2016

Mn 55 Upper, S, EEE	Mn	55
Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127904SD WG574001-10
Sample Date/Time: Wednesday, June 29, 2016 17:21:31
Number of Replicates: 3
Autosampler Position: 230
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	96610.6	3.3				ug/L	74699	Standard
	Be	9	1288.4	4.7	**1.1394**	0.049	4.3	ug/L	8	Standard
	Al	27	1461113.0	0.9	**7.6780**	0.187	2.4	ug/L	1493	Standard
	Sc	45	62711.5	6.5				ug/L	61029	Standard
	Ti	47	920.7	1.4	**2.1467**	0.013	0.6	ug/L	35	Standard
	V	51	13378.9	3.6	**1.4502**	0.085	5.9	ug/L	1435	Standard
	Cr	52	18691.9	1.7	**1.3185**	0.044	3.3	ug/L	7635	Standard
	Cr	53	1716.8	3.7	**1.3959**	0.091	6.5	ug/L	363	Standard
	Mn	55	207917.4	1.7	**27.5865**	0.619	2.2	ug/L	747	Standard
	Co	59	9044.0	2.3	**1.2346**	0.031	2.5	ug/L	234	Standard
	Ni	60	2824.3	2.7	**1.2771**	0.044	3.5	ug/L	215	Standard
	Cu	65	2601.9	3.8	**1.1945**	0.050	4.2	ug/L	104	Standard
	Zn	66	4831.1	2.2	**3.6537**	0.142	3.9	ug/L	363	Standard
>	Ge	72	533228.6	1.6				ug/L	495142	Standard
	As	75	1337.8	3.9	**1.1692**	0.061	5.2	ug/L	-50	Standard
	Se	82	140.1	2.8	**1.0475**	0.027	2.6	ug/L	14	Standard
	Se-1	77	133.7	3.0	**1.0773**	0.056	5.2	ug/L	42	Standard
>	Ga	71	393.3	11.4				mg/L	42	Standard
	Rb	85	3227.0	1.4				ug/L	17	Standard
	Y	89	450037.4	1.9				ug/L	417098	Standard
>	Rh	103	18.3	87.7				ug/L	13	Standard
	Mo	98	147.9	13.7	**0.0463**	0.009	18.8	ug/L	11	Standard
	Ag	107	7379.1	1.6	**1.0451**	0.031	3.0	ug/L	68	Standard
	Cd	111	2296.2	2.5	**1.0190**	0.058	5.7	mg/L	2	Standard
	Cd	114	5492.1	3.4	**0.9981**	0.024	2.4	ug/L	12	Standard
>	In	115	629018.5	3.2				ug/L	587441	Standard
	Sn	118	580.0	12.2	**-0.0020**	0.010	499.8	ug/L	497	Standard
	Sb	123	3693.7	7.1	**0.7349**	0.053	7.2	ug/L	78	Standard
	Ba	135	5980.2	1.6	**2.7136**	0.046	1.7	ug/L	24	Standard
	Ce	140	12665.3	1.3				ug/L	25	Standard
>	Tb	159	1007879.7	3.3				ug/L	939701	Standard
	Ho	165	265.0	7.5				ug/L	10	Standard
	Tl	203	9323.2	2.1	**1.0256**	0.019	1.8	ug/L	4	Standard
	Tl	205	8567.4	1.0	**1.0883**	0.033	3.0	ug/L	2	Standard
	Pb	206	7313.4	2.1	**1.1727**	0.042	3.6	ug/L	240	Standard
	Pb	207	6050.9	2.1	**1.0885**	0.043	4.0	ug/L	198	Standard
	Pb	208	24824.3	2.0	**1.1284**	0.015	1.4	ug/L	883	Standard
	U	238	8436.0	2.0	**1.1088**	0.013	1.2	ug/L	3	Standard
>	Bi	209	577014.5	3.2				ug/L	523984	Standard

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:23:48
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	200.0	10.9	**0.1869**	0.033	17.5	mg/L	37	Standard
K	39	61.7	46.1	**0.0210**	0.066	315.1	mg/L	38	Standard
Ca	43	46.7	12.4	**0.3502**	0.375	107.1	mg/L	10	Standard
Fe	54	378.5	13.7	**0.1751**	0.049	28.1	mg/L	179	Standard
Fe	57	365.0	16.4	**0.1419**	0.250	175.9	mg/L	335	Standard
> **Sc-1**	45	62711.5	6.5				mg/L	61029	Standard
Cl	35	27760.1	3.1				ug/L	20195	Standard
Kr	83	3.3	45.8				ug/L	4	Standard
Br	81	2080.1	7.5				ug/L	1547	Standard
P	31	18322.7	2.7				ug/L	15286	Standard
S	34	1913.5	2.4				ug/L	1572	Standard
Sr	88	98.3	10.6				ug/L	73	Standard
C	12	103.3	31.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	615.8	1.7				mg/L	6	Standard
Ho-1	165	265.0	7.5				mg/L	10	Standard
Er	166	296.7	10.3				mg/L	13	Standard
I	127	5109.3	32.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		129.334	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		107.692	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127904SD WG574001-10
Report Date/Time: Wednesday, June 29, 2016 17:23:48
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		107.078
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.121
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127905
Sample Date/Time: Wednesday, June 29, 2016 17:24:42
Number of Replicates: 3
Autosampler Position: 231
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	98923.6	1.9				ug/L	74699	Standard
	Be	9	425.0	5.1	**0.3722**	0.020	5.4	ug/L	8	Standard
	Al	27	40729181.9	0.9	**209.2967**	3.040	1.5	ug/L	1493	Standard
	Sc	45	74965.2	4.5				ug/L	61029	Standard
	Ti	47	31002.7	2.7	**78.7018**	1.731	2.2	ug/L	35	Standard
	V	51	100570.7	0.4	**12.6354**	0.094	0.7	ug/L	1435	Standard
	Cr	52	71317.5	0.6	**8.2486**	0.021	0.3	ug/L	7635	Standard
	Cr	53	8882.6	3.6	**8.9299**	0.295	3.3	ug/L	363	Standard
	Mn	55	1183907.3	1.1	**163.7014**	1.986	1.2	ug/L	747	Standard
	Co	59	16826.7	0.9	**2.4180**	0.026	1.1	ug/L	234	Standard
	Ni	60	14706.8	1.9	**7.4444**	0.110	1.5	ug/L	215	Standard
	Cu	65	7822.7	1.1	**3.9461**	0.031	0.8	ug/L	104	Standard
	Zn	66	28811.4	1.3	**24.9594**	0.201	0.8	ug/L	363	Standard
>	**Ge**	72	513872.0	0.5				ug/L	495142	Standard
	As	75	2302.8	2.2	**2.0406**	0.054	2.6	ug/L	-50	Standard
	Se	82	647.6	3.3	**5.5477**	0.199	3.6	ug/L	14	Standard
	Se-1	77	492.7	1.9	**5.8772**	0.106	1.8	ug/L	42	Standard
>	**Ga**	71	9201.1	2.2				mg/L	42	Standard
	Rb	85	118130.4	0.8				ug/L	17	Standard
	Y	89	544220.1	0.8				ug/L	417098	Standard
>	**Rh**	103	83.3	19.3				ug/L	13	Standard
	Mo	98	4972.1	0.8	**1.6995**	0.011	0.6	ug/L	11	Standard
	Ag	107	221.0	6.7	**0.0181**	0.002	11.5	ug/L	68	Standard
	Cd	111	99.4	11.4	**0.0355**	0.005	13.9	mg/L	2	Standard
	Cd	114	258.8	15.4	**0.0440**	0.007	16.4	ug/L	12	Standard
>	**In**	115	642627.7	0.7				ug/L	587441	Standard
	Sn	118	4954.1	4.3	**0.6806**	0.035	5.1	ug/L	497	Standard
	Sb	123	697.4	10.3	**0.1324**	0.015	11.0	ug/L	78	Standard
	Ba	135	175907.0	0.9	**78.5481**	0.493	0.6	ug/L	24	Standard
	Ce	140	308696.3	0.2				ug/L	25	Standard
>	**Tb**	159	1018138.8	1.9				ug/L	939701	Standard
	Ho	165	5902.8	0.8				ug/L	10	Standard
	Tl	203	694.0	1.8	**0.0782**	0.000	0.4	ug/L	4	Standard
	Tl	205	615.0	15.7	**0.0817**	0.011	14.1	ug/L	2	Standard
	Pb	206	15714.8	1.1	**2.7233**	0.017	0.6	ug/L	240	Standard
	Pb	207	12858.1	1.8	**2.5034**	0.026	1.0	ug/L	198	Standard
	Pb	208	54664.4	0.6	**2.6954**	0.031	1.2	ug/L	883	Standard
	U	238	8998.7	0.4	**1.2376**	0.021	1.7	ug/L	3	Standard
>	**Bi**	209	551643.9	1.7				ug/L	523984	Standard

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:26:59
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**46.3980**	45.060	97.1	mg/L	0	Standard
Mg	24	7338.4	2.5	**7.6542**	0.303	4.0	mg/L	37	Standard
K	39	668.3	6.0	**1.0889**	0.064	5.9	mg/L	38	Standard
Ca	43	676.7	13.4	**32.4506**	6.121	18.9	mg/L	10	Standard
Fe	54	6512.7	2.4	**4.3601**	0.118	2.7	mg/L	179	Standard
Fe	57	2150.2	3.4	**4.8854**	0.383	7.8	mg/L	335	Standard
Sc-1	45	74965.2	4.5				mg/L	61029	Standard
Cl	35	28544.2	1.7				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	11494.4	5.7				ug/L	1547	Standard
P	31	34679.6	1.6				ug/L	15286	Standard
S	34	2073.5	2.2				ug/L	1572	Standard
Sr	88	356.7	8.9				ug/L	73	Standard
C	12	163.3	55.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	3	Standard
Dy	164	9835.4	4.4				mg/L	6	Standard
Ho-1	165	5902.8	0.8				mg/L	10	Standard
Er	166	5037.5	9.2				mg/L	13	Standard
I	127	62917.2	4.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		132.430	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		103.783	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:26:59
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.394
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.279
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:26:59
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606127905
Sample Date/Time: Wednesday, June 29, 2016 17:27:54
Number of Replicates: 3
Autosampler Position: 232
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95902.7	2.3				ug/L	74699	Standard
	Be	9	26.7	54.1	**0.0312**	0.013	41.0	ug/L	8	Standard
	Al	27	866639.8	0.6	**4.5819**	0.130	2.8	ug/L	1493	Standard
	Sc	45	64717.9	1.3				ug/L	61029	Standard
	Ti	47	665.3	5.2	**1.4927**	0.099	6.6	ug/L	35	Standard
	V	51	3898.6	6.9	**0.2752**	0.030	11.1	ug/L	1435	Standard
	Cr	52	9682.4	1.6	**0.1679**	0.023	13.9	ug/L	7635	Standard
	Cr	53	670.0	10.3	**0.3316**	0.072	21.7	ug/L	363	Standard
	Mn	55	29584.3	8.9	**3.7268**	0.331	8.9	ug/L	747	Standard
	Co	59	652.3	12.4	**0.0528**	0.011	20.6	ug/L	234	Standard
	Ni	60	547.0	3.7	**0.1426**	0.010	7.1	ug/L	215	Standard
	Cu	65	341.3	5.6	**0.0636**	0.011	17.0	ug/L	104	Standard
	Zn	66	2764.3	3.0	**1.8532**	0.047	2.5	ug/L	363	Standard
>	**Ge**	72	542890.6	1.0				ug/L	495142	Standard
	As	75	11.1	320.4	**0.0688**	0.029	42.5	ug/L	-50	Standard
	Se	82	31.5	11.6	**0.1237**	0.032	26.0	ug/L	14	Standard
	Se-1	77	57.7	24.7	**0.0971**	0.173	178.1	ug/L	42	Standard
>	**Ga**	71	253.3	16.8				mg/L	42	Standard
	Rb	85	2388.5	2.4				ug/L	17	Standard
	Y	89	480185.0	1.7				ug/L	417098	Standard
>	**Rh**	103	16.7	34.6				ug/L	13	Standard
	Mo	98	144.5	13.4	**0.0432**	0.007	17.2	ug/L	11	Standard
	Ag	107	273.0	44.6	**0.0250**	0.018	70.2	ug/L	68	Standard
	Cd	111	17.1	70.4	**-0.0005**	0.005	1058.7	mg/L	2	Standard
	Cd	114	57.8	101.1	**0.0081**	0.010	129.1	ug/L	12	Standard
>	**In**	115	652419.9	1.9				ug/L	587441	Standard
	Sn	118	705.0	11.1	**0.0141**	0.014	98.8	ug/L	497	Standard
	Sb	123	247.5	32.3	**0.0437**	0.016	37.3	ug/L	78	Standard
	Ba	135	3842.5	3.8	**1.6750**	0.079	4.7	ug/L	24	Standard
	Ce	140	6778.2	3.5				ug/L	25	Standard
>	**Tb**	159	1017174.6	0.9				ug/L	939701	Standard
	Ho	165	125.0	20.8				ug/L	10	Standard
	Tl	203	149.7	88.3	**0.0148**	0.015	98.5	ug/L	4	Standard
	Tl	205	133.3	80.1	**0.0172**	0.014	79.0	ug/L	2	Standard
	Pb	206	748.0	16.8	**0.0568**	0.021	37.3	ug/L	240	Standard
	Pb	207	613.7	10.2	**0.0482**	0.012	24.5	ug/L	198	Standard
	Pb	208	2588.8	11.0	**0.0521**	0.014	26.1	ug/L	883	Standard
	U	238	255.3	28.9	**0.0295**	0.010	33.0	ug/L	3	Standard
>	**Bi**	209	573070.1	0.3				ug/L	523984	Standard

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:30:11
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.8927**	30.982	173.2	mg/L	0	Standard
Mg	24	208.3	5.5	**0.1886**	0.017	8.9	mg/L	37	Standard
K	39	35.0	14.3	**-0.0411**	0.011	27.7	mg/L	38	Standard
Ca	43	31.7	18.2	**-0.6344**	0.363	57.2	mg/L	10	Standard
Fe	54	281.0	17.0	**0.0865**	0.040	45.9	mg/L	179	Standard
Fe	57	258.3	4.5	**-0.2461**	0.039	15.8	mg/L	335	Standard
> **Sc-1**	45	64717.9	1.3				mg/L	61029	Standard
Cl	35	29516.1	1.1				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2206.8	1.5				ug/L	1547	Standard
P	31	18893.4	3.4				ug/L	15286	Standard
S	34	2140.2	7.5				ug/L	1572	Standard
Sr	88	85.0	23.5				ug/L	73	Standard
C	12	63.3	39.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	226.9	17.7				mg/L	6	Standard
Ho-1	165	125.0	20.8				mg/L	10	Standard
Er	166	133.3	38.5				mg/L	13	Standard
I	127	4113.9	27.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		128.386	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.643	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606127905
Report Date/Time: Wednesday, June 29, 2016 17:30:11
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.061
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.368
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606128101
Sample Date/Time: Wednesday, June 29, 2016 17:31:05
Number of Replicates: 3
Autosampler Position: 233
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95248.6	2.8				ug/L	74699	Standard
	Be	9	20.0	25.0	**0.0254**	0.004	15.8	ug/L	8	Standard
	Al	27	13040665.4	1.7	**69.5953**	1.139	1.6	ug/L	1493	Standard
	Sc	45	60164.2	2.0				ug/L	61029	Standard
	Ti	47	192.7	16.8	**0.4120**	0.087	21.1	ug/L	35	Standard
	V	51	8357.2	4.3	**0.9387**	0.048	5.1	ug/L	1435	Standard
	Cr	52	21860.8	2.2	**2.0003**	0.082	4.1	ug/L	7635	Standard
	Cr	53	28106.1	2.4	**30.8448**	0.828	2.7	ug/L	363	Standard
	Mn	55	35106.1	13.3	**5.0193**	0.674	13.4	ug/L	747	Standard
	Co	59	6553.8	6.0	**0.9799**	0.056	5.7	ug/L	234	Standard
	Ni	60	21271.9	1.7	**11.5228**	0.221	1.9	ug/L	215	Standard
	Cu	65	394730.4	1.5	**217.3087**	4.048	1.9	ug/L	104	Standard
	Zn	66	2800.3	3.3	**2.1735**	0.095	4.3	ug/L	363	Standard
>	**Ge**	72	483031.9	0.6				ug/L	495142	Standard
	As	75	2746.8	5.9	**2.5735**	0.152	5.9	ug/L	-50	Standard
	Se	82	892.7	0.7	**8.1998**	0.096	1.2	ug/L	14	Standard
	Se-1	77	2323.2	2.5	**31.9871**	0.998	3.1	ug/L	42	Standard
>	**Ga**	71	476.7	6.4				mg/L	42	Standard
	Rb	85	346107.5	1.7				ug/L	17	Standard
	Y	89	415017.7	1.3				ug/L	417098	Standard
>	**Rh**	103	195.0	26.0				ug/L	13	Standard
	Mo	98	2991.9	3.1	**1.1870**	0.038	3.2	ug/L	11	Standard
	Ag	107	283.3	10.3	**0.0333**	0.004	13.4	ug/L	68	Standard
	Cd	111	47.5	106.4	**0.0162**	0.025	157.7	mg/L	2	Standard
	Cd	114	106.7	47.6	**0.0200**	0.010	52.0	ug/L	12	Standard
>	**In**	115	552878.2	0.7				ug/L	587441	Standard
	Sn	118	3292.0	3.2	**0.5041**	0.021	4.2	ug/L	497	Standard
	Sb	123	2817.9	0.9	**0.6371**	0.010	1.5	ug/L	78	Standard
	Ba	135	115236.2	0.6	**59.8088**	0.708	1.2	ug/L	24	Standard
	Ce	140	265.0	80.7				ug/L	25	Standard
>	**Tb**	159	913633.5	1.5				ug/L	939701	Standard
	Ho	165	16.7	45.8				ug/L	10	Standard
	Tl	203	132.0	58.6	**0.0163**	0.011	65.5	ug/L	4	Standard
	Tl	205	136.7	39.5	**0.0217**	0.009	39.6	ug/L	2	Standard
	Pb	206	460.3	26.2	**0.0258**	0.026	100.0	ug/L	240	Standard
	Pb	207	366.3	38.5	**0.0169**	0.034	199.7	ug/L	198	Standard
	Pb	208	1427.4	18.1	**0.0116**	0.016	136.8	ug/L	883	Standard
	U	238	88.7	86.2	**0.0101**	0.013	124.4	ug/L	3	Standard
>	**Bi**	209	465790.3	0.8				ug/L	523984	Standard

Sample ID: L1606128101
Report Date/Time: Wednesday, June 29, 2016 17:33:22
Page 1

Approved: July 04, 2016

Na	23	8.3	69.3	**97.9562**	69.038	70.5	mg/L	0	Standard
Mg	24	359246.1	3.7	**470.2454**	9.358	2.0	mg/L	37	Standard
K	39	24917.1	2.4	**55.7677**	1.861	3.3	mg/L	38	Standard
Ca	43	1330.1	6.9	**82.7516**	6.716	8.1	mg/L	10	Standard
Fe	54	200.7	20.7	**0.0339**	0.032	94.4	mg/L	179	Standard
Fe	57	1015.0	10.0	**2.4309**	0.415	17.1	mg/L	335	Standard
> **Sc-1**	45	60164.2	2.0				mg/L	61029	Standard
Cl	35	38169.4	0.5				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	223802.3	3.7				ug/L	1547	Standard
P	31	18462.9	1.8				ug/L	15286	Standard
S	34	2156.8	3.7				ug/L	1572	Standard
Sr	88	483.3	16.2				ug/L	73	Standard
C	12	163.3	15.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	24.7				mg/L	3	Standard
Dy	164	32.2	122.0				mg/L	6	Standard
Ho-1	165	16.7	45.8				mg/L	10	Standard
Er	166	23.3	99.0				mg/L	13	Standard
I	127	949274.2	9.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		127.510	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.554	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606128101
Report Date/Time: Wednesday, June 29, 2016 17:33:22
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		94.116
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		88.894
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Cu 65 Upper, S, EEE	Cu	65	

Sample ID: L1606128101
Report Date/Time: Wednesday, June 29, 2016 17:33:22
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 17:34:19
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	89879.3	3.0				ug/L	74699	Standard
	Be	9	54419.2	2.1	**51.4171**	2.152	4.2	ug/L	8	Standard
	Al	27	8256861.3	1.5	**46.7135**	1.756	3.8	ug/L	1493	Standard
	Sc	45	65086.2	2.0				ug/L	61029	Standard
	Ti	47	41554.9	1.2	**98.5544**	1.167	1.2	ug/L	35	Standard
	V	51	424002.3	1.3	**50.3253**	0.743	1.5	ug/L	1435	Standard
	Cr	52	412948.6	1.6	**49.1234**	1.130	2.3	ug/L	7635	Standard
	Cr	53	52872.1	1.4	**51.1473**	0.096	0.2	ug/L	363	Standard
	Mn	55	374015.3	1.0	**48.1957**	0.700	1.5	ug/L	747	Standard
	Co	59	377049.8	1.0	**51.3472**	0.802	1.6	ug/L	234	Standard
	Ni	60	101943.5	0.6	**48.8763**	0.675	1.4	ug/L	215	Standard
	Cu	65	101706.9	0.4	**49.0730**	0.490	1.0	ug/L	104	Standard
	Zn	66	60100.6	1.2	**49.0556**	0.165	0.3	ug/L	363	Standard
>	Ge	72	550256.1	1.3				ug/L	495142	Standard
	As	75	60702.3	2.0	**48.8234**	0.606	1.2	ug/L	-50	Standard
	Se	82	6059.9	1.5	**49.5453**	0.269	0.5	ug/L	14	Standard
	Se-1	77	4264.3	0.7	**51.9185**	0.539	1.0	ug/L	42	Standard
>	Ga	71	80.0	16.5				mg/L	42	Standard
	Rb	85	943.4	9.9				ug/L	17	Standard
	Y	89	456697.4	1.9				ug/L	417098	Standard
>	Rh	103	31.7	32.9				ug/L	13	Standard
	Mo	98	286840.3	1.0	**98.7855**	2.014	2.0	ug/L	11	Standard
	Ag	107	350242.5	1.1	**49.3132**	0.224	0.5	ug/L	68	Standard
	Cd	111	110565.3	0.7	**48.5306**	0.244	0.5	mg/L	2	Standard
	Cd	114	270182.5	3.7	**48.3546**	1.830	3.8	ug/L	12	Standard
>	In	115	640022.5	1.2				ug/L	587441	Standard
	Sn	118	309740.9	1.1	**48.5773**	0.779	1.6	ug/L	497	Standard
	Sb	123	247251.7	1.0	**48.6061**	0.525	1.1	ug/L	78	Standard
	Ba	135	103942.0	1.5	**46.5980**	0.695	1.5	ug/L	24	Standard
	Ce	140	81.7	18.7				ug/L	25	Standard
>	Tb	159	1011086.0	0.8				ug/L	939701	Standard
	Ho	165	13.3	94.4				ug/L	10	Standard
	Tl	203	438780.7	1.5	**49.1134**	0.474	1.0	ug/L	4	Standard
	Tl	205	398617.2	2.3	**51.4045**	0.522	1.0	ug/L	2	Standard
	Pb	206	284136.3	1.1	**49.0127**	0.824	1.7	ug/L	240	Standard
	Pb	207	252190.2	0.9	**48.9661**	0.976	2.0	ug/L	198	Standard
	Pb	208	1023481.0	1.1	**50.2926**	0.682	1.4	ug/L	883	Standard
	U	238	380510.7	2.1	**50.9985**	0.305	0.6	ug/L	3	Standard
>	Bi	209	567945.0	2.2				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:36:36
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.3524**	31.779	173.2	mg/L	0	Standard
Mg	24	4450.7	2.4	**5.3249**	0.232	4.3	mg/L	37	Standard
K	39	2040.1	4.0	**4.1161**	0.223	5.4	mg/L	38	Standard
Ca	43	101.7	19.9	**3.5019**	1.219	34.8	mg/L	10	Standard
Fe	54	6522.9	1.9	**5.0481**	0.108	2.1	mg/L	179	Standard
Fe	57	1916.8	5.3	**5.0325**	0.229	4.6	mg/L	335	Standard
Sc-1	45	65086.2	2.0				mg/L	61029	Standard
Cl	35	28870.1	1.5				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	3490.4	22.1				ug/L	1547	Standard
P	31	20260.2	3.0				ug/L	15286	Standard
S	34	2420.2	2.6				ug/L	1572	Standard
Sr	88	90.0	19.2				ug/L	73	Standard
C	12	86.7	6.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	9.5	5.0				mg/L	6	Standard
Ho-1	165	13.3	94.4				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	23970.8	64.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	102.834		
Al	27	93.427		
Sc	45			
Ti	47	98.554		
V	51	100.651		
Cr	52	98.247		
Cr	53			
Mn	55	96.391		
Co	59	102.694		
Ni	60	97.753		
Cu	65	98.146		
Zn	66	98.111		
Ge	72		111.131	
As	75	97.647		
Se	82	99.091		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:36:36
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	98.786		
Ag	107	98.626		
Cd	111	97.061		
Cd	114			
> In	115		108.951	
Sn	118	97.155		
Sb	123	97.212		
Ba	135	93.196		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	98.227		
Tl	205			
Pb	206			
Pb	207			
Pb	208	100.585		
U	238	101.997		
> Bi	209		108.390	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:36:36
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 17:37:30
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90192.8	3.2				ug/L	74699	Standard
	Be	9	55.0	102.4	**0.0584**	0.051	86.9	ug/L	8	Standard
	Al	27	41697.2	123.7	**0.2160**	0.280	129.5	ug/L	1493	Standard
	Sc	45	66865.6	0.3				ug/L	61029	Standard
	Ti	47	70.3	63.0	**0.0618**	0.107	173.0	ug/L	35	Standard
	V	51	1545.5	1.7	**-0.0068**	0.004	62.6	ug/L	1435	Standard
	Cr	52	7551.9	5.0	**-0.0837**	0.044	52.6	ug/L	7635	Standard
	Cr	53	1056.7	15.9	**0.7243**	0.162	22.4	ug/L	363	Standard
	Mn	55	2100.5	33.7	**0.1299**	0.092	71.2	ug/L	747	Standard
	Co	59	307.3	14.8	**0.0056**	0.006	108.6	ug/L	234	Standard
	Ni	60	288.7	41.6	**0.0182**	0.059	321.7	ug/L	215	Standard
	Cu	65	326.7	64.3	**0.0580**	0.103	178.3	ug/L	104	Standard
	Zn	66	465.0	14.2	**-0.0620**	0.054	86.6	ug/L	363	Standard
>	**Ge**	72	536712.4	0.6				ug/L	495142	Standard
	As	75	-14.2	293.5	**0.0478**	0.034	71.9	ug/L	-50	Standard
	Se	82	26.6	3.7	**0.0859**	0.008	9.9	ug/L	14	Standard
	Se-1	77	71.3	25.4	**0.2781**	0.225	81.0	ug/L	42	Standard
>	**Ga**	71	51.7	22.3				mg/L	42	Standard
	Rb	85	111.7	73.6				ug/L	17	Standard
	Y	89	476031.3	1.1				ug/L	417098	Standard
>	**Rh**	103	26.7	47.2				ug/L	13	Standard
	Mo	98	310.3	31.4	**0.0963**	0.033	33.9	ug/L	11	Standard
	Ag	107	196.3	41.5	**0.0135**	0.011	82.4	ug/L	68	Standard
	Cd	111	44.4	76.9	**0.0108**	0.014	134.5	mg/L	2	Standard
	Cd	114	106.2	73.2	**0.0161**	0.013	83.2	ug/L	12	Standard
>	**In**	115	671001.8	0.7				ug/L	587441	Standard
	Sn	118	950.0	11.1	**0.0476**	0.015	31.2	ug/L	497	Standard
	Sb	123	463.6	17.0	**0.0827**	0.014	17.2	ug/L	78	Standard
	Ba	135	113.3	53.3	**0.0326**	0.026	80.4	ug/L	24	Standard
	Ce	140	100.0	52.9				ug/L	25	Standard
>	**Tb**	159	1042844.6	0.4				ug/L	939701	Standard
	Ho	165	21.7	53.3				ug/L	10	Standard
	Tl	203	169.7	59.1	**0.0166**	0.011	64.7	ug/L	4	Standard
	Tl	205	136.7	25.7	**0.0173**	0.004	24.8	ug/L	2	Standard
	Pb	206	350.0	21.1	**-0.0125**	0.012	95.5	ug/L	240	Standard
	Pb	207	292.7	21.7	**-0.0148**	0.012	78.6	ug/L	198	Standard
	Pb	208	1272.0	10.4	**-0.0133**	0.006	44.8	ug/L	883	Standard
	U	238	124.3	70.3	**0.0117**	0.011	96.0	ug/L	3	Standard
>	**Bi**	209	584508.5	0.7				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:39:47
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	165.0	89.4	**0.1290**	0.173	134.2	mg/L	37	Standard
K	39	35.0	14.3	**-0.0435**	0.010	22.9	mg/L	38	Standard
Ca	43	33.3	22.9	**-0.6014**	0.442	73.5	mg/L	10	Standard
Fe	54	222.3	16.4	**0.0337**	0.028	82.1	mg/L	179	Standard
Fe	57	293.3	1.0	**-0.1642**	0.011	6.9	mg/L	335	Standard
> **Sc-1**	45	66865.6	0.3				mg/L	61029	Standard
Cl	35	28993.0	1.1				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	2536.9	15.9				ug/L	1547	Standard
P	31	21593.7	3.8				ug/L	15286	Standard
S	34	2441.9	4.5				ug/L	1572	Standard
Sr	88	71.7	39.7				ug/L	73	Standard
C	12	43.3	53.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	2.2	286.4				mg/L	6	Standard
Ho-1	165	21.7	53.3				mg/L	10	Standard
Er	166	23.3	65.5				mg/L	13	Standard
I	127	6326.4	22.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.396	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:39:47
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.225
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		111.551
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 17:39:47
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129403
Sample Date/Time: Wednesday, June 29, 2016 17:40:43
Number of Replicates: 3
Autosampler Position: 234
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	91138.4	2.1				ug/L	74699	Standard
	Be	9	18.3	31.5	**0.0246**	0.005	20.1	ug/L	8	Standard
	Al	27	19782.9	2.5	**0.0970**	0.003	3.0	ug/L	1493	Standard
	Sc	45	69032.0	3.0				ug/L	61029	Standard
	Ti	47	54.0	3.2	**0.0194**	0.002	12.6	ug/L	35	Standard
	V	51	1966.3	6.5	**0.0390**	0.011	28.1	ug/L	1435	Standard
	Cr	52	10805.9	1.1	**0.2915**	0.043	14.7	ug/L	7635	Standard
	Cr	53	2143.5	1.8	**1.7623**	0.077	4.4	ug/L	363	Standard
	Mn	55	3844.5	2.2	**0.3500**	0.018	5.0	ug/L	747	Standard
	Co	59	327.3	8.7	**0.0074**	0.003	41.0	ug/L	234	Standard
	Ni	60	356.0	5.4	**0.0475**	0.008	16.5	ug/L	215	Standard
	Cu	65	696.0	2.6	**0.2337**	0.014	6.0	ug/L	104	Standard
	Zn	66	2205.8	3.0	**1.3661**	0.024	1.8	ug/L	363	Standard
>	**Ge**	72	549152.6	2.2				ug/L	495142	Standard
	As	75	-29.6	63.4	**0.0356**	0.015	42.9	ug/L	-50	Standard
	Se	82	23.9	20.2	**0.0586**	0.040	68.9	ug/L	14	Standard
	Se-1	77	94.7	7.0	**0.5478**	0.105	19.2	ug/L	42	Standard
>	**Ga**	71	51.7	14.8				mg/L	42	Standard
	Rb	85	138.3	25.6				ug/L	17	Standard
	Y	89	465979.4	3.6				ug/L	417098	Standard
>	**Rh**	103	13.3	21.7				ug/L	13	Standard
	Mo	98	5139.1	1.4	**1.7220**	0.056	3.3	ug/L	11	Standard
	Ag	107	110.0	6.9	**0.0022**	0.001	48.4	ug/L	68	Standard
	Cd	111	15.0	26.6	**-0.0015**	0.002	117.9	mg/L	2	Standard
	Cd	114	55.2	20.3	**0.0075**	0.002	25.2	ug/L	12	Standard
>	**In**	115	655841.8	1.8				ug/L	587441	Standard
	Sn	118	1048.4	10.6	**0.0659**	0.015	22.6	ug/L	497	Standard
	Sb	123	219.8	14.7	**0.0379**	0.005	14.5	ug/L	78	Standard
	Ba	135	307.7	5.8	**0.1187**	0.008	7.0	ug/L	24	Standard
	Ce	140	120.0	27.3				ug/L	25	Standard
>	**Tb**	159	1025647.2	2.3				ug/L	939701	Standard
	Ho	165	10.0	50.0				ug/L	10	Standard
	Tl	203	102.7	26.3	**0.0095**	0.003	29.3	ug/L	4	Standard
	Tl	205	93.3	17.2	**0.0120**	0.002	17.6	ug/L	2	Standard
	Pb	206	348.0	5.5	**-0.0123**	0.003	25.6	ug/L	240	Standard
	Pb	207	291.0	7.0	**-0.0146**	0.003	21.1	ug/L	198	Standard
	Pb	208	1220.0	1.5	**-0.0152**	0.002	10.4	ug/L	883	Standard
	U	238	43.7	3.5	**0.0013**	0.000	20.2	ug/L	3	Standard
>	**Bi**	209	579311.4	1.7				ug/L	523984	Standard

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:43:00
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	70.0	14.3	**0.0151**	0.014	91.8	mg/L	37	Standard
K	39	41.7	36.7	**-0.0323**	0.032	98.5	mg/L	38	Standard
Ca	43	23.3	24.7	**-1.2220**	0.307	25.1	mg/L	10	Standard
Fe	54	180.7	18.1	**-0.0033**	0.021	642.1	mg/L	179	Standard
Fe	57	228.3	18.9	**-0.3899**	0.112	28.8	mg/L	335	Standard
Sc-1	45	69032.0	3.0				mg/L	61029	Standard
Cl	35	29393.1	0.1				ug/L	20195	Standard
Kr	83	3.7	41.7				ug/L	4	Standard
Br	81	2280.2	16.1				ug/L	1547	Standard
P	31	20889.4	5.8				ug/L	15286	Standard
S	34	2440.2	2.2				ug/L	1572	Standard
Sr	88	78.3	9.8				ug/L	73	Standard
C	12	90.0	19.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	5.2	126.6				mg/L	6	Standard
Ho-1	165	10.0	50.0				mg/L	10	Standard
Er	166	30.0	66.7				mg/L	13	Standard
I	127	4447.3	11.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		122.008	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		110.908	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:43:00
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.644
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.559
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129403
Sample Date/Time: Wednesday, June 29, 2016 17:43:55
Number of Replicates: 3
Autosampler Position: 235
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89753.4	1.3				ug/L	74699	Standard
	Be	9	13.3	43.3	**0.0203**	0.006	27.7	ug/L	8	Standard
	Al	27	7761551.3	2.6	**43.9444**	1.176	2.7	ug/L	1493	Standard
	Sc	45	64855.2	1.4				ug/L	61029	Standard
	Ti	47	206.0	13.1	**0.3915**	0.063	16.1	ug/L	35	Standard
	V	51	6651.8	1.7	**0.6159**	0.016	2.5	ug/L	1435	Standard
	Cr	52	10534.3	3.3	**0.2860**	0.045	15.8	ug/L	7635	Standard
	Cr	53	2358.5	8.3	**2.0219**	0.206	10.2	ug/L	363	Standard
	Mn	55	48108.9	1.2	**6.2182**	0.039	0.6	ug/L	747	Standard
	Co	59	779.0	1.7	**0.0714**	0.002	2.2	ug/L	234	Standard
	Ni	60	2066.1	2.9	**0.8928**	0.027	3.0	ug/L	215	Standard
	Cu	65	1897.5	0.8	**0.8356**	0.011	1.4	ug/L	104	Standard
	Zn	66	4161.9	4.0	**3.0555**	0.122	4.0	ug/L	363	Standard
>	**Ge**	72	537411.3	0.6				ug/L	495142	Standard
	As	75	1930.1	2.9	**1.6472**	0.048	2.9	ug/L	-50	Standard
	Se	82	65.9	12.2	**0.4151**	0.071	17.1	ug/L	14	Standard
	Se-1	77	143.0	7.4	**1.1820**	0.141	12.0	ug/L	42	Standard
>	**Ga**	71	281.7	11.8				mg/L	42	Standard
	Rb	85	9322.9	0.6				ug/L	17	Standard
	Y	89	483828.6	0.3				ug/L	417098	Standard
>	**Rh**	103	65.0	35.3				ug/L	13	Standard
	Mo	98	2420.2	2.2	**0.8025**	0.008	1.0	ug/L	11	Standard
	Ag	107	86.7	9.0	**-0.0011**	0.001	98.3	ug/L	68	Standard
	Cd	111	21.1	19.4	**0.0011**	0.002	172.0	mg/L	2	Standard
	Cd	114	57.2	8.0	**0.0078**	0.001	8.8	ug/L	12	Standard
>	**In**	115	659947.0	1.4				ug/L	587441	Standard
	Sn	118	1036.7	2.5	**0.0632**	0.002	2.8	ug/L	497	Standard
	Sb	123	514.2	4.3	**0.0939**	0.005	5.0	ug/L	78	Standard
	Ba	135	42292.6	1.8	**18.3786**	0.354	1.9	ug/L	24	Standard
	Ce	140	906.7	5.0				ug/L	25	Standard
>	**Tb**	159	1004743.5	1.0				ug/L	939701	Standard
	Ho	165	30.0	33.3				ug/L	10	Standard
	Tl	203	321.7	3.2	**0.0348**	0.001	3.5	ug/L	4	Standard
	Tl	205	263.3	19.0	**0.0346**	0.006	17.3	ug/L	2	Standard
	Pb	206	528.0	6.8	**0.0212**	0.005	24.5	ug/L	240	Standard
	Pb	207	478.3	8.2	**0.0244**	0.009	35.8	ug/L	198	Standard
	Pb	208	1973.7	3.1	**0.0244**	0.005	18.6	ug/L	883	Standard
	U	238	903.7	2.1	**0.1184**	0.002	1.8	ug/L	3	Standard
>	**Bi**	209	560080.8	1.7				ug/L	523984	Standard

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:46:12
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	9985.0	3.9	**12.0618**	0.357	3.0	mg/L	37	Standard
K	39	420.0	7.4	**0.7593**	0.061	8.1	mg/L	38	Standard
Ca	43	263.3	2.9	**13.1350**	0.654	5.0	mg/L	10	Standard
Fe	54	210.7	11.6	**0.0297**	0.017	57.3	mg/L	179	Standard
Fe	57	393.3	9.5	**0.1838**	0.119	64.8	mg/L	335	Standard
> **Sc-1**	45	64855.2	1.4				mg/L	61029	Standard
Cl	35	32877.0	2.8				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	8902.6	4.1				ug/L	1547	Standard
P	31	19829.6	4.1				ug/L	15286	Standard
S	34	2240.2	4.2				ug/L	1572	Standard
Sr	88	221.7	14.7				ug/L	73	Standard
C	12	126.7	4.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	64.1	36.5				mg/L	6	Standard
Ho-1	165	30.0	33.3				mg/L	10	Standard
Er	166	53.3	21.7				mg/L	13	Standard
I	127	24610.0	6.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		120.154	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		108.537	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606129403
Report Date/Time: Wednesday, June 29, 2016 17:46:12
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.343
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.889
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129601
Sample Date/Time: Wednesday, June 29, 2016 17:47:06
Number of Replicates: 3
Autosampler Position: 236
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89414.7	0.8				ug/L	74699	Standard
	Be	9	18.3	41.7	**0.0250**	0.007	28.7	ug/L	8	Standard
	Al	27	159581.9	4.0	**0.8936**	0.029	3.3	ug/L	1493	Standard
	Sc	45	67088.2	1.1				ug/L	61029	Standard
	Ti	47	35.7	1.6	**-0.0268**	0.002	7.8	ug/L	35	Standard
	V	51	2172.9	6.2	**0.0559**	0.017	29.8	ug/L	1435	Standard
	Cr	52	9358.6	1.2	**0.0800**	0.026	32.1	ug/L	7635	Standard
	Cr	53	761.7	7.6	**0.3908**	0.063	16.2	ug/L	363	Standard
	Mn	55	2132.8	0.4	**0.1197**	0.005	4.4	ug/L	747	Standard
	Co	59	232.7	3.6	**-0.0065**	0.001	22.5	ug/L	234	Standard
	Ni	60	263.3	7.3	**-0.0009**	0.011	1213.2	ug/L	215	Standard
	Cu	65	215.7	8.8	**-0.0024**	0.009	395.1	ug/L	104	Standard
	Zn	66	2535.5	4.0	**1.5735**	0.060	3.8	ug/L	363	Standard
>	**Ge**	72	566628.0	1.5				ug/L	495142	Standard
	As	75	0.7	2838.5	**0.0602**	0.015	24.5	ug/L	-50	Standard
	Se	82	16.5	13.6	**-0.0062**	0.020	325.1	ug/L	14	Standard
	Se-1	77	57.7	30.4	**0.0671**	0.205	305.9	ug/L	42	Standard
>	**Ga**	71	65.0	35.3				mg/L	42	Standard
	Rb	85	218.3	8.7				ug/L	17	Standard
	Y	89	474347.0	1.6				ug/L	417098	Standard
>	**Rh**	103	16.7	45.8				ug/L	13	Standard
	Mo	98	108.8	9.6	**0.0309**	0.003	10.7	ug/L	11	Standard
	Ag	107	170.0	11.0	**0.0105**	0.003	26.1	ug/L	68	Standard
	Cd	111	6.5	32.0	**-0.0051**	0.001	17.0	mg/L	2	Standard
	Cd	114	10.6	76.1	**-0.0002**	0.001	576.3	ug/L	12	Standard
>	**In**	115	655393.1	1.0				ug/L	587441	Standard
	Sn	118	536.7	6.3	**-0.0123**	0.006	49.5	ug/L	497	Standard
	Sb	123	179.2	10.0	**0.0302**	0.004	12.2	ug/L	78	Standard
	Ba	135	901.0	1.9	**0.3786**	0.012	3.1	ug/L	24	Standard
	Ce	140	38.3	52.7				ug/L	25	Standard
>	**Tb**	159	1019773.4	1.2				ug/L	939701	Standard
	Ho	165	10.0	50.0				ug/L	10	Standard
	Tl	203	34.7	13.6	**0.0021**	0.001	27.9	ug/L	4	Standard
	Tl	205	33.3	52.7	**0.0044**	0.002	51.3	ug/L	2	Standard
	Pb	206	333.0	5.0	**-0.0144**	0.003	23.5	ug/L	240	Standard
	Pb	207	282.3	6.2	**-0.0157**	0.004	25.6	ug/L	198	Standard
	Pb	208	1187.0	2.0	**-0.0164**	0.002	10.6	ug/L	883	Standard
	U	238	27.0	25.9	**-0.0009**	0.001	108.4	ug/L	3	Standard
>	**Bi**	209	574894.3	1.4				ug/L	523984	Standard

Sample ID: L1606129601
Report Date/Time: Wednesday, June 29, 2016 17:49:23
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.3542**	30.050	173.2	mg/L	0	Standard
Mg	24	255.0	18.0	**0.2345**	0.054	23.2	mg/L	37	Standard
K	39	41.7	36.7	**-0.0306**	0.030	98.0	mg/L	38	Standard
Ca	43	31.7	18.2	**-0.7022**	0.343	48.8	mg/L	10	Standard
Fe	54	182.6	4.3	**0.0025**	0.004	177.8	mg/L	179	Standard
Fe	57	238.3	19.9	**-0.3363**	0.155	46.0	mg/L	335	Standard
Sc-1	45	67088.2	1.1				mg/L	61029	Standard
Cl	35	32679.2	0.8				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	2083.5	4.4				ug/L	1547	Standard
P	31	19699.5	2.2				ug/L	15286	Standard
S	34	2248.5	3.2				ug/L	1572	Standard
Sr	88	70.0	56.7				ug/L	73	Standard
C	12	60.0	76.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	42.4	37.5				mg/L	6	Standard
Ho-1	165	10.0	50.0				mg/L	10	Standard
Er	166	20.0	86.6				mg/L	13	Standard
I	127	3452.1	27.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		119.700	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.437	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606129601
Report Date/Time: Wednesday, June 29, 2016 17:49:23
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.567
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.716
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Wednesday, June 29, 2016 17:50:19
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	88962.1	0.9				ug/L	74699	Standard
	Be	9	13.3	43.3	**0.0204**	0.006	27.5	ug/L	8	Standard
	Al	27	8687475.6	2.0	**49.6312**	1.401	2.8	ug/L	1493	Standard
	Sc	45	67439.8	1.2				ug/L	61029	Standard
	Ti	47	49495.1	1.8	**118.6714**	2.776	2.3	ug/L	35	Standard
	V	51	2497.1	11.5	**0.1050**	0.033	31.3	ug/L	1435	Standard
	Cr	52	8893.6	0.7	**0.0677**	0.012	17.4	ug/L	7635	Standard
	Cr	53	8300.6	3.8	**7.8359**	0.175	2.2	ug/L	363	Standard
	Mn	55	1506.4	1.9	**0.0487**	0.003	6.1	ug/L	747	Standard
	Co	59	411.3	2.2	**0.0194**	0.002	11.4	ug/L	234	Standard
	Ni	60	986.0	7.4	**0.3548**	0.029	8.2	ug/L	215	Standard
	Cu	65	754.7	3.5	**0.2650**	0.009	3.4	ug/L	104	Standard
	Zn	66	1997.5	3.6	**1.2085**	0.053	4.4	ug/L	363	Standard
>	Ge	72	544457.7	1.7				ug/L	495142	Standard
	As	75	-43.8	164.6	**0.0238**	0.058	245.3	ug/L	-50	Standard
	Se	82	18.8	19.8	**0.0181**	0.034	186.0	ug/L	14	Standard
	Se-1	77	455.0	4.4	**5.0437**	0.245	4.9	ug/L	42	Standard
>	Ga	71	186.7	17.4				mg/L	42	Standard
	Rb	85	1130.0	9.7				ug/L	17	Standard
	Y	89	464215.3	1.7				ug/L	417098	Standard
>	Rh	103	10.0	100.0				ug/L	13	Standard
	Mo	98	300673.2	0.9	**102.5347**	1.457	1.4	ug/L	11	Standard
	Ag	107	252.3	6.3	**0.0223**	0.003	12.6	ug/L	68	Standard
	Cd	111	-194.8	14.0	**-0.0926**	0.012	12.9	mg/L	2	Standard
	Cd	114	577.1	3.2	**0.1002**	0.004	4.2	ug/L	12	Standard
>	In	115	646355.5	1.8				ug/L	587441	Standard
	Sn	118	791.7	5.4	**0.0285**	0.006	21.5	ug/L	497	Standard
	Sb	123	195.6	12.5	**0.0338**	0.004	11.9	ug/L	78	Standard
	Ba	135	712.0	4.0	**0.3001**	0.008	2.7	ug/L	24	Standard
	Ce	140	246.7	7.1				ug/L	25	Standard
>	Tb	159	1024202.6	1.1				ug/L	939701	Standard
	Ho	165	23.3	61.9				ug/L	10	Standard
	Tl	203	122.3	9.8	**0.0119**	0.001	11.8	ug/L	4	Standard
	Tl	205	90.0	11.1	**0.0118**	0.001	10.5	ug/L	2	Standard
	Pb	206	533.7	2.0	**0.0208**	0.002	10.5	ug/L	240	Standard
	Pb	207	414.7	3.2	**0.0105**	0.002	21.4	ug/L	198	Standard
	Pb	208	1800.4	3.2	**0.0144**	0.003	21.4	ug/L	883	Standard
	U	238	18.7	20.3	**-0.0019**	0.001	26.8	ug/L	3	Standard
>	Bi	209	568876.5	0.5				ug/L	523984	Standard

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 17:52:36
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	11507.7	5.3	**13.3841**	0.868	6.5	mg/L	37	Standard
K	39	2163.5	5.0	**4.2129**	0.203	4.8	mg/L	38	Standard
Ca	43	273.3	20.1	**13.1286**	3.349	25.5	mg/L	10	Standard
Fe	54	17120.6	2.5	**12.9991**	0.385	3.0	mg/L	179	Standard
Fe	57	4470.7	3.8	**12.6745**	0.387	3.1	mg/L	335	Standard
> **Sc-1**	45	67439.8	1.2				mg/L	61029	Standard
Cl	35	28777.3	2.0				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	2086.8	10.9				ug/L	1547	Standard
P	31	20233.5	1.2				ug/L	15286	Standard
S	34	2398.5	7.6				ug/L	1572	Standard
Sr	88	63.3	35.6				ug/L	73	Standard
C	12	173.3	8.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	80.0	69.6				mg/L	3	Standard
Dy	164	11.6	52.9				mg/L	6	Standard
Ho-1	165	23.3	61.9				mg/L	10	Standard
Er	166	36.7	56.8				mg/L	13	Standard
I	127	2830.3	21.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27	0.993		
Sc	45			
Ti	47	118.671		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.960	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 17:52:36
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	102.535		
Ag	107			
Cd	111			
Cd	114			
> In	115			110.029
Sn	118			
Sb	123			
Ba	135			
Ce	140			
> Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
> Bi	209			108.567
Na	23	0.040		
Mg	24	267.683		
K	39	84.259		
Ca	43	87.524		
Fe	54	103.992		
Fe	57	101.396		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Na	23	
QC Std 4	Mg	24	

Sample ID: QC Std 4
Report Date/Time: Wednesday, June 29, 2016 17:52:36
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Wednesday, June 29, 2016 17:53:31
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88668.7	1.6				ug/L	74699	Standard
	Be	9	115505.6	3.3	**110.5278**	2.817	2.5	ug/L	8	Standard
	Al	27	8938177.8	2.5	**51.2202**	0.471	0.9	ug/L	1493	Standard
	Sc	45	65087.9	3.1				ug/L	61029	Standard
	Ti	47	50197.4	1.2	**122.5394**	1.491	1.2	ug/L	35	Standard
	V	51	859522.7	1.3	**105.1949**	1.380	1.3	ug/L	1435	Standard
	Cr	52	831098.2	2.0	**102.8267**	1.943	1.9	ug/L	7635	Standard
	Cr	53	106911.0	2.3	**106.7884**	2.096	2.0	ug/L	363	Standard
	Mn	55	780341.4	1.6	**103.6433**	1.453	1.4	ug/L	747	Standard
	Co	59	785536.0	0.5	**110.1234**	0.547	0.5	ug/L	234	Standard
	Ni	60	203753.2	2.0	**100.6522**	1.808	1.8	ug/L	215	Standard
	Cu	65	201732.1	0.8	**100.2703**	0.833	0.8	ug/L	104	Standard
	Zn	66	122540.9	0.7	**103.4323**	0.376	0.4	ug/L	363	Standard
>	**Ge**	72	534677.6	0.4				ug/L	495142	Standard
	As	75	124456.8	0.2	**102.9556**	0.389	0.4	ug/L	-50	Standard
	Se	82	12596.7	0.4	**106.1488**	0.433	0.4	ug/L	14	Standard
	Se-1	77	8794.6	0.8	**110.8893**	1.382	1.2	ug/L	42	Standard
>	**Ga**	71	163.3	10.8				mg/L	42	Standard
	Rb	85	966.7	4.8				ug/L	17	Standard
	Y	89	476525.6	1.7				ug/L	417098	Standard
>	**Rh**	103	56.7	53.9				ug/L	13	Standard
	Mo	98	309227.9	0.6	**104.8818**	1.132	1.1	ug/L	11	Standard
	Ag	107	592523.9	2.1	**82.1821**	1.932	2.4	ug/L	68	Standard
	Cd	111	228625.7	1.0	**98.8467**	0.904	0.9	mg/L	2	Standard
	Cd	114	551600.1	1.5	**97.2335**	1.523	1.6	ug/L	12	Standard
>	**In**	115	649797.8	0.6				ug/L	587441	Standard
	Sn	118	1576.7	4.2	**0.1493**	0.011	7.5	ug/L	497	Standard
	Sb	123	495027.3	1.9	**95.8585**	2.358	2.5	ug/L	78	Standard
	Ba	135	210763.3	1.3	**93.0830**	1.687	1.8	ug/L	24	Standard
	Ce	140	70.0	24.7				ug/L	25	Standard
>	**Tb**	159	1005734.4	0.5				ug/L	939701	Standard
	Ho	165	11.7	49.5				ug/L	10	Standard
	Tl	203	893894.1	1.2	**101.2389**	0.317	0.3	ug/L	4	Standard
	Tl	205	815034.2	0.6	**106.3608**	0.654	0.6	ug/L	2	Standard
	Pb	206	583528.7	0.9	**101.9143**	0.243	0.2	ug/L	240	Standard
	Pb	207	499626.1	1.2	**98.2110**	0.372	0.4	ug/L	198	Standard
	Pb	208	2105185.2	0.8	**104.7451**	0.431	0.4	ug/L	883	Standard
	U	238	779429.7	1.7	**105.7078**	0.634	0.6	ug/L	3	Standard
>	**Bi**	209	561247.5	1.2				ug/L	523984	Standard

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.1880**	31.494	173.2	mg/L	0	Standard
Mg	24	12009.8	3.1	**14.4710**	0.020	0.1	mg/L	37	Standard
K	39	2150.2	4.2	**4.3456**	0.256	5.9	mg/L	38	Standard
Ca	43	251.7	12.1	**12.4083**	2.087	16.8	mg/L	10	Standard
Fe	54	17836.2	2.6	**14.0504**	0.592	4.2	mg/L	179	Standard
Fe	57	4537.3	4.4	**13.3981**	0.845	6.3	mg/L	335	Standard
> **Sc-1**	45	65087.9	3.1				mg/L	61029	Standard
Cl	35	27559.7	2.9				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	1986.8	4.8				ug/L	1547	Standard
P	31	19052.0	2.9				ug/L	15286	Standard
S	34	2286.8	6.3				ug/L	1572	Standard
Sr	88	50.0	10.0				ug/L	73	Standard
C	12	120.0	54.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	6.0	200.7				mg/L	6	Standard
Ho-1	165	11.7	49.5				mg/L	10	Standard
Er	166	13.3	86.6				mg/L	13	Standard
I	127	2065.1	25.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	110.528		
Al	27	1.024		
Sc	45			
Ti	47	122.539		
V	51	105.195		
Cr	52	102.827		
Cr	53			
Mn	55	103.643		
Co	59	110.123		
Ni	60	100.652		
Cu	65	100.270		
Zn	66	103.432		
> Ge	72		107.985	
As	75	102.956		
Se	82	106.149		
Se-1	77			
> Ga	71			

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	104.882		
Ag	107	82.182		
Cd	111	98.847		
Cd	114			
In	115		110.615	
Sn	118			
Sb	123	95.858		
Ba	135	93.083		
Ce	140			
Tb	159			
Ho	165			
Tl	203	101.239		
Tl	205			
Pb	206			
Pb	207			
Pb	208	104.745		
U	238	105.708		
Bi	209		107.111	
Na	23	145.504		
Mg	24	289.419		
K	39	86.911		
Ca	43	82.722		
Fe	54	112.403		
Fe	57	107.185		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Ti	47	
QC Std 5	Na	23	

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 3

Approved: July 04, 2016

QC Std 5	Mg	24

Sample ID: QC Std 5
Report Date/Time: Wednesday, June 29, 2016 17:55:47
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 17:56:44
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88001.7	3.8				ug/L	74699	Standard
	Be	9	58491.1	3.4	**56.4121**	0.807	1.4	ug/L	8	Standard
	Al	27	8741223.1	1.9	**50.5126**	1.685	3.3	ug/L	1493	Standard
	Sc	45	66566.0	2.9				ug/L	61029	Standard
	Ti	47	44065.3	0.6	**105.8464**	2.393	2.3	ug/L	35	Standard
	V	51	439728.9	0.9	**52.8674**	1.388	2.6	ug/L	1435	Standard
	Cr	52	434693.1	0.1	**52.4281**	0.919	1.8	ug/L	7635	Standard
	Cr	53	53735.1	1.4	**52.6541**	1.206	2.3	ug/L	363	Standard
	Mn	55	394255.2	1.0	**51.4522**	0.719	1.4	ug/L	747	Standard
	Co	59	403494.7	0.9	**55.6456**	1.136	2.0	ug/L	234	Standard
	Ni	60	106493.3	0.2	**51.7076**	0.770	1.5	ug/L	215	Standard
	Cu	65	105229.8	0.3	**51.4208**	1.011	2.0	ug/L	104	Standard
	Zn	66	61955.4	1.5	**51.2322**	1.156	2.3	ug/L	363	Standard
>	**Ge**	72	543443.2	1.7				ug/L	495142	Standard
	As	75	62757.8	1.3	**51.1208**	1.290	2.5	ug/L	-50	Standard
	Se	82	6411.0	0.6	**53.0945**	1.065	2.0	ug/L	14	Standard
	Se-1	77	4401.6	1.7	**54.2995**	1.443	2.7	ug/L	42	Standard
>	**Ga**	71	61.7	12.4				mg/L	42	Standard
	Rb	85	943.4	4.4				ug/L	17	Standard
	Y	89	472814.1	2.0				ug/L	417098	Standard
>	**Rh**	103	31.7	36.5				ug/L	13	Standard
	Mo	98	301859.8	0.5	**101.5547**	1.508	1.5	ug/L	11	Standard
	Ag	107	366041.5	1.2	**50.3642**	1.561	3.1	ug/L	68	Standard
	Cd	111	117544.4	1.3	**50.4048**	0.748	1.5	mg/L	2	Standard
	Cd	114	286950.2	2.0	**50.1688**	0.955	1.9	ug/L	12	Standard
>	**In**	115	655193.6	1.9				ug/L	587441	Standard
	Sn	118	333959.0	1.2	**51.1818**	1.510	3.0	ug/L	497	Standard
	Sb	123	261510.0	1.1	**50.2357**	1.515	3.0	ug/L	78	Standard
	Ba	135	110572.3	0.3	**48.4318**	0.852	1.8	ug/L	24	Standard
	Ce	140	81.7	46.0				ug/L	25	Standard
>	**Tb**	159	1009652.9	1.9				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	466012.5	0.6	**52.5145**	1.216	2.3	ug/L	4	Standard
	Tl	205	428260.1	0.9	**55.6109**	1.642	3.0	ug/L	2	Standard
	Pb	206	294247.5	0.8	**51.0977**	1.194	2.3	ug/L	240	Standard
	Pb	207	260542.8	0.7	**50.9223**	1.007	2.0	ug/L	198	Standard
	Pb	208	1081629.9	0.8	**53.5152**	1.532	2.9	ug/L	883	Standard
	U	238	403586.7	0.9	**54.4635**	1.402	2.6	ug/L	3	Standard
>	**Bi**	209	564239.9	2.1				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:59:01
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4549.0	2.0	**5.3196**	0.066	1.2	mg/L	37	Standard
K	39	2091.8	2.6	**4.1250**	0.021	0.5	mg/L	38	Standard
Ca	43	106.7	9.8	**3.6589**	0.645	17.6	mg/L	10	Standard
Fe	54	6853.2	4.1	**5.1952**	0.357	6.9	mg/L	179	Standard
Fe	57	1901.8	4.2	**4.8507**	0.081	1.7	mg/L	335	Standard
> **Sc-1**	45	66566.0	2.9				mg/L	61029	Standard
Cl	35	28481.4	2.6				ug/L	20195	Standard
Kr	83	3.3	75.5				ug/L	4	Standard
Br	81	1840.1	18.3				ug/L	1547	Standard
P	31	20427.1	5.4				ug/L	15286	Standard
S	34	2438.5	6.3				ug/L	1572	Standard
Sr	88	80.0	32.5				ug/L	73	Standard
C	12	76.7	7.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	6.5	86.7				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	2150.2	26.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	112.824		
Al	27	101.025		
Sc	45			
Ti	47	105.846		
V	51	105.735		
Cr	52	104.856		
Cr	53			
Mn	55	102.904		
Co	59	111.291		
Ni	60	103.415		
Cu	65	102.842		
Zn	66	102.464		
> Ge	72		109.755	
As	75	102.242		
Se	82	106.189		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:59:01
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	101.555		
Ag	107	100.728		
Cd	111	100.810		
Cd	114			
> In	115			111.533
Sn	118	102.364		
Sb	123	100.471		
Ba	135	96.864		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	105.029		
Tl	205			
Pb	206			
Pb	207			
Pb	208	107.030		
U	238	108.927		
> Bi	209			107.683
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 17:59:01
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 17:59:55
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86353.7	1.4				ug/L	74699	Standard
	Be	9	83.3	67.4	**0.0891**	0.054	60.5	ug/L	8	Standard
	Al	27	42873.5	92.7	**0.2376**	0.231	97.1	ug/L	1493	Standard
	Sc	45	65305.4	1.2				ug/L	61029	Standard
	Ti	47	79.7	51.6	**0.0788**	0.097	123.4	ug/L	35	Standard
	V	51	1461.2	4.4	**-0.0225**	0.008	34.9	ug/L	1435	Standard
	Cr	52	7626.9	6.9	**-0.1029**	0.066	64.1	ug/L	7635	Standard
	Cr	53	760.0	14.8	**0.4057**	0.109	26.8	ug/L	363	Standard
	Mn	55	4546.9	137.2	**0.4388**	0.807	183.9	ug/L	747	Standard
	Co	59	626.0	93.3	**0.0478**	0.080	167.0	ug/L	234	Standard
	Ni	60	409.0	51.9	**0.0715**	0.102	142.7	ug/L	215	Standard
	Cu	65	272.7	44.7	**0.0274**	0.059	215.3	ug/L	104	Standard
	Zn	66	514.0	23.7	**-0.0335**	0.100	300.2	ug/L	363	Standard
>	**Ge**	72	553485.6	1.3				ug/L	495142	Standard
	As	75	-6.7	474.8	**0.0545**	0.025	45.9	ug/L	-50	Standard
	Se	82	20.4	10.3	**0.0278**	0.015	55.0	ug/L	14	Standard
	Se-1	77	66.3	4.4	**0.1906**	0.046	24.0	ug/L	42	Standard
>	**Ga**	71	50.0	10.0				mg/L	42	Standard
	Rb	85	45.0	48.4				ug/L	17	Standard
	Y	89	474090.9	1.7				ug/L	417098	Standard
>	**Rh**	103	18.3	41.7				ug/L	13	Standard
	Mo	98	533.3	54.8	**0.1739**	0.097	56.0	ug/L	11	Standard
	Ag	107	205.7	25.6	**0.0154**	0.007	45.9	ug/L	68	Standard
	Cd	111	43.3	79.3	**0.0106**	0.015	137.7	mg/L	2	Standard
	Cd	114	160.8	66.0	**0.0260**	0.018	70.8	ug/L	12	Standard
>	**In**	115	653140.9	0.6				ug/L	587441	Standard
	Sn	118	840.0	19.7	**0.0347**	0.026	74.7	ug/L	497	Standard
	Sb	123	1054.2	24.1	**0.1991**	0.050	25.2	ug/L	78	Standard
	Ba	135	82.0	54.8	**0.0199**	0.020	97.9	ug/L	24	Standard
	Ce	140	78.3	62.0				ug/L	25	Standard
>	**Tb**	159	1021325.7	0.5				ug/L	939701	Standard
	Ho	165	20.0	66.1				ug/L	10	Standard
	Tl	203	163.0	50.8	**0.0163**	0.009	56.1	ug/L	4	Standard
	Tl	205	543.4	156.5	**0.0696**	0.109	156.0	ug/L	2	Standard
	Pb	206	336.7	16.7	**-0.0136**	0.010	70.5	ug/L	240	Standard
	Pb	207	309.7	12.8	**-0.0104**	0.008	73.2	ug/L	198	Standard
	Pb	208	2162.1	83.2	**0.0313**	0.088	280.1	ug/L	883	Standard
	U	238	524.4	153.4	**0.0652**	0.107	163.8	ug/L	3	Standard
>	**Bi**	209	573163.8	0.1				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:02:12
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.8448**	30.899	173.2	mg/L	0	Standard
Mg	24	123.3	127.6	**0.0829**	0.188	227.1	mg/L	37	Standard
K	39	30.0	66.7	**-0.0524**	0.041	77.8	mg/L	38	Standard
Ca	43	21.7	35.3	**-1.2409**	0.467	37.7	mg/L	10	Standard
Fe	54	185.6	27.6	**0.0086**	0.039	460.9	mg/L	179	Standard
Fe	57	268.3	10.3	**-0.2224**	0.079	35.7	mg/L	335	Standard
Sc-1	45	65305.4	1.2				mg/L	61029	Standard
Cl	35	28806.0	1.7				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1900.1	3.3				ug/L	1547	Standard
P	31	20240.2	2.7				ug/L	15286	Standard
S	34	2435.2	7.1				ug/L	1572	Standard
Sr	88	65.0	20.4				ug/L	73	Standard
C	12	60.0					mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	25.5	84.0				mg/L	6	Standard
Ho-1	165	20.0	66.1				mg/L	10	Standard
Er	166	23.3	65.5				mg/L	13	Standard
I	127	1926.8	28.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		111.783	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:02:12
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.184
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.386
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Mn	55	

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:02:12
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 8
Sample Date/Time: Wednesday, June 29, 2016 18:03:08
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84590.6	2.3				ug/L	74699	Standard
	Be	9	260.0	12.0	**0.2687**	0.034	12.7	ug/L	8	Standard
	Al	27	3603.8	43.5	**0.0082**	0.009	110.1	ug/L	1493	Standard
	Sc	45	65034.3	0.9				ug/L	61029	Standard
	Ti	47	43.7	32.9	**-0.0008**	0.035	4199.9	ug/L	35	Standard
	V	51	4694.6	2.2	**0.3894**	0.008	2.0	ug/L	1435	Standard
	Cr	52	14076.2	0.6	**0.7639**	0.008	1.0	ug/L	7635	Standard
	Cr	53	1435.1	6.7	**1.1315**	0.098	8.6	ug/L	363	Standard
	Mn	55	4379.0	3.5	**0.4438**	0.017	3.9	ug/L	747	Standard
	Co	59	3208.0	4.2	**0.4198**	0.015	3.5	ug/L	234	Standard
	Ni	60	3464.1	1.9	**1.6175**	0.036	2.2	ug/L	215	Standard
	Cu	65	1802.1	7.0	**0.8073**	0.055	6.8	ug/L	104	Standard
	Zn	66	8137.9	1.9	**6.5566**	0.074	1.1	ug/L	363	Standard
>	**Ge**	72	526101.8	0.9				ug/L	495142	Standard
	As	75	520.8	10.5	**0.4970**	0.042	8.4	ug/L	-50	Standard
	Se	82	68.5	3.0	**0.4490**	0.013	3.0	ug/L	14	Standard
	Se-1	77	97.0	14.4	**0.6266**	0.169	26.9	ug/L	42	Standard
>	**Ga**	71	35.0	42.9				mg/L	42	Standard
	Rb	85	53.3	5.4				ug/L	17	Standard
	Y	89	462050.3	3.1				ug/L	417098	Standard
>	**Rh**	103	11.7	24.7				ug/L	13	Standard
	Mo	98	178.6	39.2	**0.0548**	0.024	44.0	ug/L	11	Standard
	Ag	107	2989.6	2.3	**0.4006**	0.011	2.7	ug/L	68	Standard
	Cd	111	543.3	9.6	**0.2264**	0.024	10.6	mg/L	2	Standard
	Cd	114	1493.6	11.4	**0.2605**	0.031	12.1	ug/L	12	Standard
>	**In**	115	651766.1	0.6				ug/L	587441	Standard
	Sn	118	551.7	10.5	**-0.0096**	0.009	97.6	ug/L	497	Standard
	Sb	123	2328.3	1.4	**0.4453**	0.009	1.9	ug/L	78	Standard
	Ba	135	1600.4	3.6	**0.6889**	0.030	4.3	ug/L	24	Standard
	Ce	140	6.7	86.6				ug/L	25	Standard
>	**Tb**	159	1010895.5	0.8				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	860.7	17.5	**0.0936**	0.016	16.7	ug/L	4	Standard
	Tl	205	718.4	10.6	**0.0918**	0.008	8.5	ug/L	2	Standard
	Pb	206	1533.1	9.5	**0.1909**	0.020	10.6	ug/L	240	Standard
	Pb	207	1247.1	5.1	**0.1700**	0.007	4.3	ug/L	198	Standard
	Pb	208	5357.0	3.3	**0.1870**	0.005	2.6	ug/L	883	Standard
	U	238	3040.0	2.0	**0.3994**	0.014	3.4	ug/L	3	Standard
>	**Bi**	209	573346.3	2.5				ug/L	523984	Standard

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 18:05:25
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**18.2178**	31.546	173.2	mg/L	0	Standard
Mg	24	50.0	50.0	**-0.0044**	0.031	706.6	mg/L	37	Standard
K	39	35.0	24.7	**-0.0415**	0.018	44.0	mg/L	38	Standard
Ca	43	26.7	43.3	**-0.9446**	0.676	71.5	mg/L	10	Standard
Fe	54	105.0	32.5	**-0.0546**	0.028	51.3	mg/L	179	Standard
Fe	57	230.0	24.7	**-0.3395**	0.189	55.5	mg/L	335	Standard
Sc-1	45	65034.3	0.9				mg/L	61029	Standard
Cl	35	27747.4	0.7				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	1776.8	9.7				ug/L	1547	Standard
P	31	14345.1	1.0				ug/L	15286	Standard
S	34	2316.8	2.0				ug/L	1572	Standard
Sr	88	78.3	41.5				ug/L	73	Standard
C	12	70.0	28.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	6.0	100.4				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	1788.4	30.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	134.333		
Al	27			
Sc	45			
Ti	47			
V	51	97.347		
Cr	52	95.485		
Cr	53			
Mn	55	88.762		
Co	59	104.961		
Ni	60	101.093		
Cu	65	100.908		
Zn	66	104.906		
Ge	72		106.253	
As	75	124.241		
Se	82	112.254		
Se-1	77			
Ga	71			

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 18:05:25
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	100.142		
Cd	111	94.325		
Cd	114			
> In	115			110.950
Sn	118			
Sb	123	111.327		
Ba	135	91.848		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	116.967		
Tl	205			
Pb	206			
Pb	207			
Pb	208	93.500		
U	238	99.844		
> Bi	209			109.420
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 8	Be	9	

Sample ID: QC Std 8
Report Date/Time: Wednesday, June 29, 2016 18:05:25
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBS M9 WG574192-02
Sample Date/Time: Wednesday, June 29, 2016 18:06:21
Number of Replicates: 3
Autosampler Position: 237
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84464.8	1.8				ug/L	74699	Standard
	Be	9	13.3	94.4	**0.0212**	0.013	60.4	ug/L	8	Standard
	Al	27	11484.4	1.6	**0.0557**	0.001	1.2	ug/L	1493	Standard
	Sc	45	65588.4	2.6				ug/L	61029	Standard
	Ti	47	51.3	20.8	**0.0150**	0.026	174.5	ug/L	35	Standard
	V	51	1499.5	1.3	**-0.0137**	0.003	18.5	ug/L	1435	Standard
	Cr	52	8909.3	1.2	**0.0773**	0.005	6.6	ug/L	7635	Standard
	Cr	53	853.4	5.0	**0.5155**	0.034	6.6	ug/L	363	Standard
	Mn	55	1324.4	4.8	**0.0262**	0.011	41.6	ug/L	747	Standard
	Co	59	237.0	9.4	**-0.0044**	0.004	80.7	ug/L	234	Standard
	Ni	60	317.3	8.0	**0.0315**	0.014	46.0	ug/L	215	Standard
	Cu	65	277.0	2.0	**0.0327**	0.005	14.4	ug/L	104	Standard
	Zn	66	2117.5	3.0	**1.3211**	0.055	4.2	ug/L	363	Standard
>	**Ge**	72	540634.9	1.6				ug/L	495142	Standard
	As	75	-19.3	191.5	**0.0441**	0.030	68.3	ug/L	-50	Standard
	Se	82	22.9	30.4	**0.0532**	0.060	113.4	ug/L	14	Standard
	Se-1	77	66.3	3.1	**0.2096**	0.028	13.1	ug/L	42	Standard
>	**Ga**	71	40.0	45.1				mg/L	42	Standard
	Rb	85	45.0	40.1				ug/L	17	Standard
	Y	89	495113.6	1.2				ug/L	417098	Standard
>	**Rh**	103	20.0	50.0				ug/L	13	Standard
	Mo	98	103.4	4.5	**0.0285**	0.002	7.3	ug/L	11	Standard
	Ag	107	105.3	22.1	**0.0013**	0.003	227.8	ug/L	68	Standard
	Cd	111	14.8	35.7	**-0.0017**	0.002	127.2	mg/L	2	Standard
	Cd	114	30.3	51.7	**0.0031**	0.003	86.4	ug/L	12	Standard
>	**In**	115	666604.9	1.7				ug/L	587441	Standard
	Sn	118	860.0	2.1	**0.0351**	0.004	11.7	ug/L	497	Standard
	Sb	123	400.3	27.4	**0.0716**	0.022	30.7	ug/L	78	Standard
	Ba	135	277.0	2.3	**0.1032**	0.001	0.8	ug/L	24	Standard
	Ce	140	75.0	20.0				ug/L	25	Standard
>	**Tb**	159	1002064.8	2.5				ug/L	939701	Standard
	Ho	165	13.3	21.7				ug/L	10	Standard
	Tl	203	23.7	14.8	**0.0009**	0.000	40.1	ug/L	4	Standard
	Tl	205	18.3	56.8	**0.0026**	0.001	51.0	ug/L	2	Standard
	Pb	206	383.3	7.2	**-0.0041**	0.004	93.5	ug/L	240	Standard
	Pb	207	304.7	4.5	**-0.0100**	0.002	23.2	ug/L	198	Standard
	Pb	208	1385.4	1.0	**-0.0050**	0.001	21.8	ug/L	883	Standard
	U	238	16.0	18.8	**-0.0023**	0.000	16.8	ug/L	3	Standard
>	**Bi**	209	560531.0	1.7				ug/L	523984	Standard

Sample ID: PBS M9 WG574192-02
Report Date/Time: Wednesday, June 29, 2016 18:08:38
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	51.7	11.2	**-0.0029**	0.008	286.8	mg/L	37	Standard
K	39	28.3	27.0	**-0.0560**	0.015	26.7	mg/L	38	Standard
Ca	43	25.0	34.6	**-1.0611**	0.483	45.5	mg/L	10	Standard
Fe	54	209.4	8.6	**0.0268**	0.012	44.3	mg/L	179	Standard
Fe	57	206.7	9.2	**-0.4211**	0.048	11.4	mg/L	335	Standard
> **Sc-1**	45	65588.4	2.6				mg/L	61029	Standard
Cl	35	29295.6	1.0				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	3250.3	3.7				ug/L	1547	Standard
P	31	19926.4	1.7				ug/L	15286	Standard
S	34	2431.9	7.4				ug/L	1572	Standard
Sr	88	63.3	32.9				ug/L	73	Standard
C	12	60.0	16.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	23.2	64.8				mg/L	6	Standard
Ho-1	165	13.3	21.7				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	1665.1	25.5				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		113.074	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		109.188	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: PBS M9 WG574192-02
Report Date/Time: Wednesday, June 29, 2016 18:08:38
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.476
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.975
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBS M9 WG574192-02
Report Date/Time: Wednesday, June 29, 2016 18:08:38
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSS M9 WG574192-03
Sample Date/Time: Wednesday, June 29, 2016 18:09:32
Number of Replicates: 3
Autosampler Position: 238
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83350.4	2.2				ug/L	74699	Standard
	Be	9	28323.1	1.8	**28.8492**	0.865	3.0	ug/L	8	Standard
	Al	27	10196.8	1.6	**0.0488**	0.002	4.6	ug/L	1493	Standard
	Sc	45	65183.2	0.6				ug/L	61029	Standard
	Ti	47	95.7	13.6	**0.1113**	0.027	24.3	ug/L	35	Standard
	V	51	224256.1	0.6	**25.7588**	0.522	2.0	ug/L	1435	Standard
	Cr	52	224696.6	1.0	**25.4792**	0.636	2.5	ug/L	7635	Standard
	Cr	53	27321.3	1.7	**25.5112**	0.743	2.9	ug/L	363	Standard
	Mn	55	202745.7	1.5	**25.3082**	0.750	3.0	ug/L	747	Standard
	Co	59	203564.8	2.3	**26.9126**	1.011	3.8	ug/L	234	Standard
	Ni	60	54379.7	0.7	**25.2634**	0.489	1.9	ug/L	215	Standard
	Cu	65	53975.6	1.0	**25.2465**	0.617	2.4	ug/L	104	Standard
	Zn	66	33582.8	1.1	**26.4199**	0.666	2.5	ug/L	363	Standard
>	**Ge**	72	566570.7	1.5				ug/L	495142	Standard
	As	75	31915.6	1.0	**24.9643**	0.450	1.8	ug/L	-50	Standard
	Se	82	3210.1	2.6	**25.4315**	0.960	3.8	ug/L	14	Standard
	Se-1	77	2232.8	4.1	**26.1060**	1.397	5.4	ug/L	42	Standard
>	**Ga**	71	91.7	13.7				mg/L	42	Standard
	Rb	85	96.7	31.2				ug/L	17	Standard
	Y	89	463441.3	2.1				ug/L	417098	Standard
>	**Rh**	103	25.0	40.0				ug/L	13	Standard
	Mo	98	76.3	5.4	**0.0207**	0.001	6.1	ug/L	11	Standard
	Ag	107	186996.6	1.0	**26.4732**	0.339	1.3	ug/L	68	Standard
	Cd	111	58916.9	1.6	**26.0048**	0.541	2.1	mg/L	2	Standard
	Cd	114	141808.0	1.8	**25.5238**	0.616	2.4	ug/L	12	Standard
>	**In**	115	636399.6	0.6				ug/L	587441	Standard
	Sn	118	665.0	9.8	**0.0103**	0.010	94.1	ug/L	497	Standard
	Sb	123	127977.1	1.2	**25.3000**	0.453	1.8	ug/L	78	Standard
	Ba	135	55490.8	0.9	**25.0111**	0.350	1.4	ug/L	24	Standard
	Ce	140	61.7	36.6				ug/L	25	Standard
>	**Tb**	159	977939.2	1.1				ug/L	939701	Standard
	Ho	165	5.0	100.0				ug/L	10	Standard
	Tl	203	232757.9	0.8	**26.2994**	0.658	2.5	ug/L	4	Standard
	Tl	205	213539.2	0.1	**27.8001**	0.582	2.1	ug/L	2	Standard
	Pb	206	150898.4	0.7	**26.2407**	0.649	2.5	ug/L	240	Standard
	Pb	207	129115.2	1.2	**25.2720**	0.800	3.2	ug/L	198	Standard
	Pb	208	547777.7	0.6	**27.1359**	0.615	2.3	ug/L	883	Standard
	U	238	197609.3	0.3	**26.7341**	0.481	1.8	ug/L	3	Standard
>	**Bi**	209	562718.9	2.0				ug/L	523984	Standard

Sample ID: LCSS M9 WG574192-03
Report Date/Time: Wednesday, June 29, 2016 18:11:49
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**35.9815**	31.158	86.6	mg/L	0	Standard
Mg	24	53.3	23.6	**-0.0007**	0.015	2138.9	mg/L	37	Standard
K	39	36.7	90.8	**-0.0380**	0.070	183.0	mg/L	38	Standard
Ca	43	26.7	28.6	**-0.9444**	0.462	48.9	mg/L	10	Standard
Fe	54	149.2	16.7	**-0.0198**	0.020	102.0	mg/L	179	Standard
Fe	57	258.3	10.7	**-0.2518**	0.092	36.6	mg/L	335	Standard
> **Sc-1**	45	65183.2	0.6				mg/L	61029	Standard
Cl	35	28068.6	2.1				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	1853.4	3.5				ug/L	1547	Standard
P	31	20221.8	1.7				ug/L	15286	Standard
S	34	2256.8	4.8				ug/L	1572	Standard
Sr	88	75.0	24.0				ug/L	73	Standard
C	12	113.3	53.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	2.9	203.0				mg/L	6	Standard
Ho-1	165	5.0	100.0				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	1621.8	26.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.582	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.426	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSS M9 WG574192-03
Report Date/Time: Wednesday, June 29, 2016 18:11:49
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.334
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.392
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSS M9 WG574192-03
Report Date/Time: Wednesday, June 29, 2016 18:11:49
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120601 WG574192-01
Sample Date/Time: Wednesday, June 29, 2016 18:12:44
Number of Replicates: 3
Autosampler Position: 239
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89463.4	1.2				ug/L	74699	Standard
	Be	9	386.7	13.6	**0.3740**	0.046	12.2	ug/L	8	Standard
	Al	27	7605562.2	1.6	**43.1976**	0.340	0.8	ug/L	1493	Standard
	Sc	45	72243.5	1.6				ug/L	61029	Standard
	Ti	47	9362.6	0.8	**21.8169**	0.679	3.1	ug/L	35	Standard
	V	51	123334.5	1.4	**14.2946**	0.165	1.2	ug/L	1435	Standard
	Cr	52	107034.1	1.6	**11.7966**	0.478	4.1	ug/L	7635	Standard
	Cr	53	12683.6	2.9	**11.8556**	0.642	5.4	ug/L	363	Standard
	Mn	55	2726083.5	2.1	**347.2409**	2.026	0.6	ug/L	747	Standard
	Co	59	29531.4	3.0	**3.9299**	0.043	1.1	ug/L	234	Standard
	Ni	60	23196.4	0.4	**10.8725**	0.312	2.9	ug/L	215	Standard
	Cu	65	77156.3	0.7	**36.6848**	0.692	1.9	ug/L	104	Standard
	Zn	66	748003.6	1.6	**607.1548**	7.539	1.2	ug/L	363	Standard
>	**Ge**	72	558122.4	2.5				ug/L	495142	Standard
	As	75	3312.8	2.4	**2.6835**	0.016	0.6	ug/L	-50	Standard
	Se	82	112.5	6.7	**0.7726**	0.083	10.8	ug/L	14	Standard
	Se-1	77	130.7	4.7	**0.9656**	0.087	9.0	ug/L	42	Standard
>	**Ga**	71	7793.7	3.0				mg/L	42	Standard
	Rb	85	97459.1	1.1				ug/L	17	Standard
	Y	89	585281.5	2.1				ug/L	417098	Standard
>	**Rh**	103	56.7	13.5				ug/L	13	Standard
	Mo	98	1826.8	4.5	**0.5887**	0.034	5.8	ug/L	11	Standard
	Ag	107	1189.0	7.2	**0.1453**	0.012	8.4	ug/L	68	Standard
	Cd	111	5739.0	1.3	**2.3714**	0.004	0.2	mg/L	2	Standard
	Cd	114	13674.5	3.2	**2.3090**	0.056	2.4	ug/L	12	Standard
>	**In**	115	677670.9	1.2				ug/L	587441	Standard
	Sn	118	2238.5	6.9	**0.2377**	0.027	11.3	ug/L	497	Standard
	Sb	123	1060.8	22.9	**0.1924**	0.043	22.3	ug/L	78	Standard
	Ba	135	275486.3	0.7	**116.6679**	1.279	1.1	ug/L	24	Standard
	Ce	140	521396.3	0.5				ug/L	25	Standard
>	**Tb**	159	1067665.8	1.0				ug/L	939701	Standard
	Ho	165	8012.1	4.7				ug/L	10	Standard
	Tl	203	658.7	5.9	**0.0698**	0.004	5.8	ug/L	4	Standard
	Tl	205	588.3	7.1	**0.0738**	0.005	7.2	ug/L	2	Standard
	Pb	206	280687.1	0.1	**46.9821**	0.294	0.6	ug/L	240	Standard
	Pb	207	217657.6	0.6	**40.9978**	0.453	1.1	ug/L	198	Standard
	Pb	208	937818.9	1.0	**44.7144**	0.652	1.5	ug/L	883	Standard
	U	238	5450.3	0.5	**0.7046**	0.003	0.4	ug/L	3	Standard
>	**Bi**	209	585148.7	0.6				ug/L	523984	Standard

Sample ID: L1606120601 WG574192-01
Report Date/Time: Wednesday, June 29, 2016 18:15:01
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	733.4	5.7	**0.7343**	0.035	4.8	mg/L	37	Standard
K	39	401.7	1.9	**0.6360**	0.019	3.0	mg/L	38	Standard
Ca	43	643.3	17.5	**31.7962**	5.845	18.4	mg/L	10	Standard
Fe	54	8818.1	2.4	**6.1777**	0.130	2.1	mg/L	179	Standard
Fe	57	2673.6	4.7	**6.5999**	0.243	3.7	mg/L	335	Standard
Sc-1	45	72243.5	1.6				mg/L	61029	Standard
Cl	35	29798.6	1.4				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	3707.1	5.8				ug/L	1547	Standard
P	31	20168.4	2.0				ug/L	15286	Standard
S	34	2373.5	6.9				ug/L	1572	Standard
Sr	88	256.7	13.8				ug/L	73	Standard
C	12	350.0	7.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	130.0	53.3				mg/L	3	Standard
Dy	164	13278.3	0.5				mg/L	6	Standard
Ho-1	165	8012.1	4.7				mg/L	10	Standard
Er	166	7608.6	3.8				mg/L	13	Standard
I	127	89948.0	2.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		119.765	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		112.720	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120601 WG574192-01
Report Date/Time: Wednesday, June 29, 2016 18:15:01
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	115.360
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	111.673
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120601 WG574192-01
Report Date/Time: Wednesday, June 29, 2016 18:15:01
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120602
Sample Date/Time: Wednesday, June 29, 2016 18:15:55
Number of Replicates: 3
Autosampler Position: 240
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88784.4	1.1				ug/L	74699	Standard
	Be	9	281.7	5.7	**0.2767**	0.013	4.7	ug/L	8	Standard
	Al	27	7005404.7	2.4	**40.0898**	0.625	1.6	ug/L	1493	Standard
	Sc	45	70594.2	3.3				ug/L	61029	Standard
	Ti	47	10155.7	1.6	**24.3492**	0.468	1.9	ug/L	35	Standard
	V	51	153110.0	0.2	**18.3086**	0.101	0.6	ug/L	1435	Standard
	Cr	52	102957.6	0.7	**11.6568**	0.092	0.8	ug/L	7635	Standard
	Cr	53	11732.9	3.0	**11.2567**	0.305	2.7	ug/L	363	Standard
	Mn	55	2052735.8	1.1	**268.9604**	2.146	0.8	ug/L	747	Standard
	Co	59	33492.3	1.4	**4.5921**	0.057	1.2	ug/L	234	Standard
	Ni	60	26642.1	0.7	**12.8642**	0.053	0.4	ug/L	215	Standard
	Cu	65	40463.3	0.7	**19.7403**	0.183	0.9	ug/L	104	Standard
	Zn	66	608360.3	0.1	**507.8962**	1.599	0.3	ug/L	363	Standard
>	**Ge**	72	542468.6	0.4				ug/L	495142	Standard
	As	75	4231.0	2.7	**3.5076**	0.102	2.9	ug/L	-50	Standard
	Se	82	256.2	4.5	**1.9932**	0.099	5.0	ug/L	14	Standard
	Se-1	77	208.7	7.7	**1.9860**	0.210	10.6	ug/L	42	Standard
>	**Ga**	71	5872.8	1.9				mg/L	42	Standard
	Rb	85	68412.5	1.5				ug/L	17	Standard
	Y	89	570393.2	0.3				ug/L	417098	Standard
>	**Rh**	103	206.7	25.4				ug/L	13	Standard
	Mo	98	11374.8	0.4	**3.8637**	0.075	1.9	ug/L	11	Standard
	Ag	107	1353.7	2.3	**0.1754**	0.001	0.8	ug/L	68	Standard
	Cd	111	5491.2	1.2	**2.3732**	0.046	1.9	mg/L	2	Standard
	Cd	114	13400.9	1.9	**2.3668**	0.045	1.9	ug/L	12	Standard
>	**In**	115	648040.3	1.6				ug/L	587441	Standard
	Sn	118	1990.1	9.0	**0.2145**	0.033	15.4	ug/L	497	Standard
	Sb	123	589.0	13.2	**0.1100**	0.013	12.1	ug/L	78	Standard
	Ba	135	226861.9	1.0	**100.4795**	2.067	2.1	ug/L	24	Standard
	Ce	140	495305.9	0.7				ug/L	25	Standard
>	**Tb**	159	1028084.3	0.7				ug/L	939701	Standard
	Ho	165	7586.9	1.7				ug/L	10	Standard
	Tl	203	574.7	1.8	**0.0640**	0.002	2.9	ug/L	4	Standard
	Tl	205	501.7	7.3	**0.0663**	0.004	6.1	ug/L	2	Standard
	Pb	206	2490204.6	1.0	**439.6443**	9.561	2.2	ug/L	240	Standard
	Pb	207	1941338.7	1.0	**385.7578**	8.533	2.2	ug/L	198	Standard
	Pb	208	8788859.7	0.3	**442.0013**	4.418	1.0	ug/L	883	Standard
	U	238	27264.2	0.8	**3.7315**	0.055	1.5	ug/L	3	Standard
>	**Bi**	209	555601.8	1.3				ug/L	523984	Standard

Sample ID: L1606120602
Report Date/Time: Wednesday, June 29, 2016 18:18:12
Page 1

Approved: July 04, 2016

Na	23	15.0	66.7	**151.5495**	103.438	68.3	mg/L	0	Standard
Mg	24	875.0	7.4	**0.9112**	0.062	6.8	mg/L	37	Standard
K	39	308.3	17.3	**0.4774**	0.118	24.7	mg/L	38	Standard
Ca	43	2528.5	1.9	**135.7080**	7.063	5.2	mg/L	10	Standard
Fe	54	8423.7	4.4	**6.0457**	0.457	7.6	mg/L	179	Standard
Fe	57	3312.0	6.3	**8.6682**	0.763	8.8	mg/L	335	Standard
> **Sc-1**	45	70594.2	3.3				mg/L	61029	Standard
Cl	35	33743.5	0.4				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	4283.9	2.2				ug/L	1547	Standard
P	31	21890.8	1.3				ug/L	15286	Standard
S	34	2663.6	6.3				ug/L	1572	Standard
Sr	88	533.3	8.7				ug/L	73	Standard
C	12	600.0	5.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	100.0	26.5				mg/L	3	Standard
Dy	164	11590.5	1.3				mg/L	6	Standard
Ho-1	165	7586.9	1.7				mg/L	10	Standard
Er	166	6698.1	4.4				mg/L	13	Standard
I	127	108507.7	4.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		118.857	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.558	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120602
Report Date/Time: Wednesday, June 29, 2016 18:18:12
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	110.316
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	106.034
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120602
Report Date/Time: Wednesday, June 29, 2016 18:18:12
Page 3

Approved: July 04, 2016

Pb 206 Upper, S, EEE	Pb	206
Pb 207 Upper, S, EEE	Pb	207
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120603S WG574192-04
Sample Date/Time: Wednesday, June 29, 2016 18:19:07
Number of Replicates: 3
Autosampler Position: 241
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86561.6	0.8				ug/L	74699	Standard
	Be	9	28455.1	4.4	**27.8947**	1.031	3.7	ug/L	8	Standard
	Al	27	11326081.2	0.4	**66.4980**	0.812	1.2	ug/L	1493	Standard
	Sc	45	74033.8	1.5				ug/L	61029	Standard
	Ti	47	13830.6	1.3	**31.7218**	0.771	2.4	ug/L	35	Standard
	V	51	404083.1	0.7	**46.4723**	0.823	1.8	ug/L	1435	Standard
	Cr	52	396378.6	0.5	**45.6257**	0.448	1.0	ug/L	7635	Standard
	Cr	53	48917.2	0.9	**45.8335**	0.539	1.2	ug/L	363	Standard
	Mn	55	3610519.8	1.3	**452.1326**	4.314	1.0	ug/L	747	Standard
	Co	59	243070.7	1.0	**32.0670**	0.356	1.1	ug/L	234	Standard
	Ni	60	87954.6	1.3	**40.8489**	0.598	1.5	ug/L	215	Standard
	Cu	65	138890.7	1.2	**64.9805**	0.614	0.9	ug/L	104	Standard
	Zn	66	982893.1	0.9	**784.3980**	13.633	1.7	ug/L	363	Standard
>	**Ge**	72	567734.3	1.2				ug/L	495142	Standard
	As	75	35264.6	1.3	**27.5210**	0.578	2.1	ug/L	-50	Standard
	Se	82	3103.1	1.8	**24.5237**	0.601	2.4	ug/L	14	Standard
	Se-1	77	2151.2	2.8	**25.0663**	0.745	3.0	ug/L	42	Standard
>	**Ga**	71	14508.6	3.6				mg/L	42	Standard
	Rb	85	180073.9	1.5				ug/L	17	Standard
	Y	89	616764.0	1.9				ug/L	417098	Standard
>	**Rh**	103	85.0	11.8				ug/L	13	Standard
	Mo	98	3441.4	2.1	**1.1615**	0.033	2.8	ug/L	11	Standard
	Ag	107	170180.4	0.7	**23.5897**	0.078	0.3	ug/L	68	Standard
	Cd	111	63817.8	1.8	**27.5852**	0.736	2.7	mg/L	2	Standard
	Cd	114	155790.6	2.3	**27.4604**	0.877	3.2	ug/L	12	Standard
>	**In**	115	649912.7	0.9				ug/L	587441	Standard
	Sn	118	2755.3	3.8	**0.3317**	0.018	5.5	ug/L	497	Standard
	Sb	123	11515.0	1.7	**2.2251**	0.028	1.3	ug/L	78	Standard
	Ba	135	420339.6	0.5	**185.6246**	2.179	1.2	ug/L	24	Standard
	Ce	140	771001.7	1.2				ug/L	25	Standard
>	**Tb**	159	1033151.8	0.5				ug/L	939701	Standard
	Ho	165	11344.2	1.5				ug/L	10	Standard
	Tl	203	220759.0	0.5	**24.6088**	0.113	0.5	ug/L	4	Standard
	Tl	205	203485.6	0.4	**26.1371**	0.140	0.5	ug/L	2	Standard
	Pb	206	518067.3	0.6	**89.0511**	0.447	0.5	ug/L	240	Standard
	Pb	207	415032.4	0.7	**80.2896**	0.537	0.7	ug/L	198	Standard
	Pb	208	1813547.0	1.2	**88.8061**	1.200	1.4	ug/L	883	Standard
	U	238	196913.8	1.4	**26.2850**	0.405	1.5	ug/L	3	Standard
>	**Bi**	209	570191.4	0.2				ug/L	523984	Standard

Sample ID: L1606120603S WG574192-04
Report Date/Time: Wednesday, June 29, 2016 18:21:24
Page 1

Approved: July 04, 2016

Na	23	6.7	43.3	**63.0881**	26.284	41.7	mg/L	0	Standard
Mg	24	761.7	8.6	**0.7448**	0.057	7.6	mg/L	37	Standard
K	39	698.3	9.3	**1.1574**	0.103	8.9	mg/L	38	Standard
Ca	43	761.7	2.5	**37.1357**	0.619	1.7	mg/L	10	Standard
Fe	54	13661.1	2.3	**9.4129**	0.352	3.7	mg/L	179	Standard
Fe	57	3945.5	4.9	**9.9839**	0.617	6.2	mg/L	335	Standard
> **Sc-1**	45	74033.8	1.5				mg/L	61029	Standard
Cl	35	29672.4	3.3				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	5120.9	2.8				ug/L	1547	Standard
P	31	21184.8	2.3				ug/L	15286	Standard
S	34	2585.2	3.6				ug/L	1572	Standard
Sr	88	278.3	7.3				ug/L	73	Standard
C	12	453.3	23.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	180.0	28.9				mg/L	3	Standard
Dy	164	18295.9	3.4				mg/L	6	Standard
Ho-1	165	11344.2	1.5				mg/L	10	Standard
Er	166	10350.2	2.2				mg/L	13	Standard
I	127	155575.8	4.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		115.881	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.661	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120603S WG574192-04
Report Date/Time: Wednesday, June 29, 2016 18:21:24
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.635
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.818
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120603S WG574192-04
Report Date/Time: Wednesday, June 29, 2016 18:21:24
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120604SD WG574192-05
Sample Date/Time: Wednesday, June 29, 2016 18:22:18
Number of Replicates: 3
Autosampler Position: 242
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88318.4	2.5				ug/L	74699	Standard
	Be	9	28910.9	0.8	**27.7911**	0.552	2.0	ug/L	8	Standard
	Al	27	11678142.6	0.8	**67.2186**	1.350	2.0	ug/L	1493	Standard
	Sc	45	75929.8	1.2				ug/L	61029	Standard
	Ti	47	15202.6	0.9	**34.3685**	0.367	1.1	ug/L	35	Standard
	V	51	408247.6	0.9	**46.2682**	0.632	1.4	ug/L	1435	Standard
	Cr	52	411599.1	1.0	**46.7151**	0.586	1.3	ug/L	7635	Standard
	Cr	53	50226.5	2.4	**46.3783**	0.843	1.8	ug/L	363	Standard
	Mn	55	4064951.2	0.1	**501.7134**	5.151	1.0	ug/L	747	Standard
	Co	59	242783.8	0.7	**31.5662**	0.473	1.5	ug/L	234	Standard
	Ni	60	92537.7	0.8	**42.3583**	0.203	0.5	ug/L	215	Standard
	Cu	65	135365.1	1.4	**62.4070**	0.241	0.4	ug/L	104	Standard
	Zn	66	1025151.4	1.0	**806.2133**	6.015	0.7	ug/L	363	Standard
>	**Ge**	72	576071.4	1.0				ug/L	495142	Standard
	As	75	35892.9	0.1	**27.6040**	0.298	1.1	ug/L	-50	Standard
	Se	82	2975.5	0.8	**23.1647**	0.155	0.7	ug/L	14	Standard
	Se-1	77	2118.1	0.7	**24.3051**	0.149	0.6	ug/L	42	Standard
>	**Ga**	71	16053.5	1.5				mg/L	42	Standard
	Rb	85	192660.5	0.7				ug/L	17	Standard
	Y	89	633758.5	0.9				ug/L	417098	Standard
>	**Rh**	103	91.7	22.7				ug/L	13	Standard
	Mo	98	2104.9	2.7	**0.6926**	0.023	3.3	ug/L	11	Standard
	Ag	107	176068.8	0.4	**23.8725**	0.248	1.0	ug/L	68	Standard
	Cd	111	64701.8	0.7	**27.3512**	0.177	0.6	mg/L	2	Standard
	Cd	114	157399.0	1.1	**27.1310**	0.176	0.6	ug/L	12	Standard
>	**In**	115	664480.0	1.2				ug/L	587441	Standard
	Sn	118	2676.9	2.8	**0.3105**	0.015	4.8	ug/L	497	Standard
	Sb	123	10363.1	2.2	**1.9584**	0.055	2.8	ug/L	78	Standard
	Ba	135	439420.6	0.3	**189.8023**	2.264	1.2	ug/L	24	Standard
	Ce	140	842883.5	1.0				ug/L	25	Standard
>	**Tb**	159	1049787.1	1.3				ug/L	939701	Standard
	Ho	165	12211.6	1.7				ug/L	10	Standard
	Tl	203	222058.4	0.3	**24.4396**	0.316	1.3	ug/L	4	Standard
	Tl	205	202828.8	1.6	**25.7185**	0.095	0.4	ug/L	2	Standard
	Pb	206	574235.7	0.3	**97.4667**	1.839	1.9	ug/L	240	Standard
	Pb	207	457211.7	0.6	**87.3408**	1.861	2.1	ug/L	198	Standard
	Pb	208	1979847.3	0.4	**95.7233**	1.171	1.2	ug/L	883	Standard
	U	238	197274.4	0.9	**26.0018**	0.641	2.5	ug/L	3	Standard
>	**Bi**	209	577596.9	1.6				ug/L	523984	Standard

Sample ID: L1606120604SD WG574192-05
Report Date/Time: Wednesday, June 29, 2016 18:24:35
Page 1

Approved: July 04, 2016

Na	23	5.0	0.0	**46.3267**	0.533	1.2	mg/L	0	Standard
Mg	24	961.7	5.5	**0.9325**	0.047	5.0	mg/L	37	Standard
K	39	738.4	15.4	**1.1991**	0.216	18.0	mg/L	38	Standard
Ca	43	858.4	11.4	**41.0973**	5.443	13.2	mg/L	10	Standard
Fe	54	16143.5	2.3	**10.8618**	0.122	1.1	mg/L	179	Standard
Fe	57	4519.0	2.4	**11.2744**	0.427	3.8	mg/L	335	Standard
> **Sc-1**	45	75929.8	1.2				mg/L	61029	Standard
Cl	35	29136.0	1.4				ug/L	20195	Standard
Kr	83	1.0	0.0				ug/L	4	Standard
Br	81	5384.3	8.1				ug/L	1547	Standard
P	31	20994.5	3.6				ug/L	15286	Standard
S	34	2526.9	1.1				ug/L	1572	Standard
Sr	88	265.0	8.2				ug/L	73	Standard
C	12	466.7	15.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	193.3	16.6				mg/L	3	Standard
Dy	164	19678.1	2.3				mg/L	6	Standard
Ho-1	165	12211.6	1.7				mg/L	10	Standard
Er	166	11454.3	2.8				mg/L	13	Standard
I	127	133210.9	5.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		118.233	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		116.345	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120604SD WG574192-05
Report Date/Time: Wednesday, June 29, 2016 18:24:35
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.114
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.232
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120604SD WG574192-05
Report Date/Time: Wednesday, June 29, 2016 18:24:35
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605
Sample Date/Time: Wednesday, June 29, 2016 18:25:30
Number of Replicates: 3
Autosampler Position: 243
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	90694.1	2.2				ug/L	74699	Standard
	Be	9	526.7	6.5	**0.5004**	0.031	6.1	ug/L	8	Standard
	Al	27	8237088.6	1.8	**46.1607**	0.953	2.1	ug/L	1493	Standard
	Sc	45	76241.4	2.3				ug/L	61029	Standard
	Ti	47	9939.3	1.0	**23.0163**	0.554	2.4	ug/L	35	Standard
	V	51	146708.3	2.5	**16.9369**	0.672	4.0	ug/L	1435	Standard
	Cr	52	214286.2	1.5	**24.4749**	0.700	2.9	ug/L	7635	Standard
	Cr	53	25651.7	1.8	**24.1415**	0.427	1.8	ug/L	363	Standard
	Mn	55	2369893.5	2.4	**300.0787**	12.076	4.0	ug/L	747	Standard
	Co	59	42091.1	2.2	**5.5844**	0.210	3.8	ug/L	234	Standard
	Ni	60	29674.7	1.4	**13.8532**	0.382	2.8	ug/L	215	Standard
	Cu	65	80108.6	1.5	**37.8555**	1.188	3.1	ug/L	104	Standard
	Zn	66	402296.4	1.3	**324.3434**	9.485	2.9	ug/L	363	Standard
>	**Ge**	72	561626.1	1.7				ug/L	495142	Standard
	As	75	4613.8	0.6	**3.6915**	0.048	1.3	ug/L	-50	Standard
	Se	82	73.3	11.4	**0.4501**	0.057	12.7	ug/L	14	Standard
	Se-1	77	107.3	9.4	**0.6736**	0.125	18.5	ug/L	42	Standard
>	**Ga**	71	10595.4	4.5				mg/L	42	Standard
	Rb	85	117987.7	1.5				ug/L	17	Standard
	Y	89	650208.5	0.3				ug/L	417098	Standard
>	**Rh**	103	85.0	38.6				ug/L	13	Standard
	Mo	98	1074.4	2.1	**0.3404**	0.008	2.3	ug/L	11	Standard
	Ag	107	1521.7	2.8	**0.1876**	0.005	2.8	ug/L	68	Standard
	Cd	111	14943.9	2.9	**6.1278**	0.124	2.0	mg/L	2	Standard
	Cd	114	36335.4	1.3	**6.0807**	0.042	0.7	ug/L	12	Standard
>	**In**	115	684218.1	1.0				ug/L	587441	Standard
	Sn	118	4600.7	0.6	**0.5815**	0.007	1.3	ug/L	497	Standard
	Sb	123	346.8	6.2	**0.0596**	0.003	5.7	ug/L	78	Standard
	Ba	135	305952.3	0.8	**128.3272**	1.108	0.9	ug/L	24	Standard
	Ce	140	926252.0	0.8				ug/L	25	Standard
>	**Tb**	159	1067500.9	0.3				ug/L	939701	Standard
	Ho	165	12695.3	1.4				ug/L	10	Standard
	Tl	203	780.0	9.1	**0.0850**	0.009	10.5	ug/L	4	Standard
	Tl	205	700.0	0.7	**0.0898**	0.001	1.1	ug/L	2	Standard
	Pb	206	317389.9	0.7	**54.3657**	1.082	2.0	ug/L	240	Standard
	Pb	207	249270.6	1.1	**48.0507**	1.100	2.3	ug/L	198	Standard
	Pb	208	1104553.9	2.1	**53.8955**	1.488	2.8	ug/L	883	Standard
	U	238	5524.0	2.1	**0.7309**	0.022	3.0	ug/L	3	Standard
>	**Bi**	209	572007.3	1.4				ug/L	523984	Standard

Sample ID: L1606120605
Report Date/Time: Wednesday, June 29, 2016 18:27:47
Page 1

Approved: July 04, 2016

Na	23	8.3	34.6	**77.0658**	27.245	35.4	mg/L	0	Standard
Mg	24	545.0	4.9	**0.4986**	0.038	7.7	mg/L	37	Standard
K	39	528.3	15.8	**0.8195**	0.134	16.4	mg/L	38	Standard
Ca	43	966.7	6.0	**46.3930**	3.592	7.7	mg/L	10	Standard
Fe	54	12851.8	4.3	**8.5861**	0.399	4.6	mg/L	179	Standard
Fe	57	3822.2	7.3	**9.3402**	1.007	10.8	mg/L	335	Standard
> **Sc-1**	45	76241.4	2.3				mg/L	61029	Standard
Cl	35	29717.8	2.2				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	3207.0	9.1				ug/L	1547	Standard
P	31	20467.1	0.8				ug/L	15286	Standard
S	34	2505.2	3.0				ug/L	1572	Standard
Sr	88	263.3	8.6				ug/L	73	Standard
C	12	196.7	5.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	123.3	46.8				mg/L	3	Standard
Dy	164	19864.2	0.8				mg/L	6	Standard
Ho-1	165	12695.3	1.4				mg/L	10	Standard
Er	166	11821.3	4.6				mg/L	13	Standard
I	127	41200.9	0.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		121.413	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.427	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120605
Report Date/Time: Wednesday, June 29, 2016 18:27:47
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.474
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.165
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606120605
Report Date/Time: Wednesday, June 29, 2016 18:27:47
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605PS WG574690-01
Sample Date/Time: Wednesday, June 29, 2016 18:28:41
Number of Replicates: 3
Autosampler Position: 244
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	85371.5	1.3				ug/L	74699	Standard
	Be	9	57838.5	0.7	**57.4989**	0.995	1.7	ug/L	8	Standard
	Al	27	8190235.3	0.9	**48.7593**	1.052	2.2	ug/L	1493	Standard
	Sc	45	74174.5	1.1				ug/L	61029	Standard
	Ti	47	9905.9	1.1	**22.7993**	0.140	0.6	ug/L	35	Standard
	V	51	615589.0	1.1	**71.2575**	1.604	2.3	ug/L	1435	Standard
	Cr	52	662720.1	2.0	**77.3523**	0.990	1.3	ug/L	7635	Standard
	Cr	53	81212.1	1.6	**76.6855**	0.228	0.3	ug/L	363	Standard
	Mn	55	2850355.8	1.2	**358.7372**	7.553	2.1	ug/L	747	Standard
	Co	59	453959.9	1.8	**60.2304**	1.747	2.9	ug/L	234	Standard
	Ni	60	139316.4	1.9	**65.0924**	0.607	0.9	ug/L	215	Standard
	Cu	65	189057.7	1.5	**88.9267**	0.405	0.5	ug/L	104	Standard
	Zn	66	476114.4	1.0	**381.6097**	5.307	1.4	ug/L	363	Standard
>	Ge	72	564919.1	1.3				ug/L	495142	Standard
	As	75	71672.0	0.5	**56.1469**	0.522	0.9	ug/L	-50	Standard
	Se	82	6681.0	0.4	**53.2203**	0.482	0.9	ug/L	14	Standard
	Se-1	77	4575.0	1.8	**54.2855**	1.161	2.1	ug/L	42	Standard
>	Ga	71	10925.6	2.3				mg/L	42	Standard
	Rb	85	119128.2	1.4				ug/L	17	Standard
	Y	89	658944.3	2.5				ug/L	417098	Standard
>	Rh	103	98.3	24.0				ug/L	13	Standard
	Mo	98	1073.3	2.3	**0.3534**	0.015	4.1	ug/L	11	Standard
	Ag	107	364831.9	0.7	**49.9049**	1.416	2.8	ug/L	68	Standard
	Cd	111	132655.1	1.4	**56.5569**	1.029	1.8	mg/L	2	Standard
	Cd	114	321468.3	0.7	**55.8946**	1.680	3.0	ug/L	12	Standard
>	In	115	659074.5	2.5				ug/L	587441	Standard
	Sn	118	1318.4	1.7	**0.1066**	0.008	7.5	ug/L	497	Standard
	Sb	123	258576.1	0.4	**49.3799**	1.200	2.4	ug/L	78	Standard
	Ba	135	409891.1	0.5	**178.5483**	4.068	2.3	ug/L	24	Standard
	Ce	140	929186.0	0.4				ug/L	25	Standard
>	Tb	159	1023915.1	1.8				ug/L	939701	Standard
	Ho	165	12935.5	3.1				ug/L	10	Standard
	Tl	203	455240.8	0.8	**52.4471**	0.837	1.6	ug/L	4	Standard
	Tl	205	420642.4	0.2	**55.8346**	0.392	0.7	ug/L	2	Standard
	Pb	206	614831.5	0.9	**109.2366**	1.843	1.7	ug/L	240	Standard
	Pb	207	503849.3	0.9	**100.7500**	1.617	1.6	ug/L	198	Standard
	Pb	208	2170185.0	1.2	**109.8441**	2.227	2.0	ug/L	883	Standard
	U	238	393858.5	1.0	**54.3381**	1.023	1.9	ug/L	3	Standard
>	Bi	209	551780.7	0.8				ug/L	523984	Standard

Sample ID: L1606120605PS WG574690-01
Report Date/Time: Wednesday, June 29, 2016 18:30:58
Page 1

Approved: July 04, 2016

Na	23	10.0	86.6	**95.2332**	83.153	87.3	mg/L	0	Standard
Mg	24	453.3	5.6	**0.4164**	0.027	6.4	mg/L	37	Standard
K	39	480.0	7.5	**0.7589**	0.067	8.8	mg/L	38	Standard
Ca	43	940.0	10.6	**46.3668**	5.650	12.2	mg/L	10	Standard
Fe	54	12349.5	1.2	**8.4763**	0.023	0.3	mg/L	179	Standard
Fe	57	3702.1	2.7	**9.2777**	0.167	1.8	mg/L	335	Standard
Sc-1	45	74174.5	1.1				mg/L	61029	Standard
Cl	35	30087.8	0.9				ug/L	20195	Standard
Kr	83	2.7	43.3				ug/L	4	Standard
Br	81	3260.4	2.1				ug/L	1547	Standard
P	31	20787.6	1.7				ug/L	15286	Standard
S	34	2500.2	2.8				ug/L	1572	Standard
Sr	88	245.0	12.4				ug/L	73	Standard
C	12	160.0	54.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	103.3	22.3				mg/L	3	Standard
Dy	164	20338.3	3.5				mg/L	6	Standard
Ho-1	165	12935.5	3.1				mg/L	10	Standard
Er	166	11817.9	3.9				mg/L	13	Standard
I	127	39775.5	3.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		114.288	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.092	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120605PS WG574690-01
Report Date/Time: Wednesday, June 29, 2016 18:30:58
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	112.194
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	105.305
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120605PS WG574690-01
Report Date/Time: Wednesday, June 29, 2016 18:30:58
Page 3

Approved: July 04, 2016

Pb 206 Upper, S, EEE	Pb	206
Pb 207 Upper, S, EEE	Pb	207
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605SDL WG574690-02
Sample Date/Time: Wednesday, June 29, 2016 18:31:53
Number of Replicates: 3
Autosampler Position: 245
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	84476.5	1.1				ug/L	74699	Standard
	Be	9	178.3	18.7	**0.1869**	0.035	18.6	ug/L	8	Standard
	Al	27	1736036.7	3.1	**10.4319**	0.283	2.7	ug/L	1493	Standard
	Sc	45	72059.3	1.7				ug/L	61029	Standard
	Ti	47	2123.5	6.6	**4.6249**	0.308	6.7	ug/L	35	Standard
	V	51	32734.7	3.8	**3.4666**	0.135	3.9	ug/L	1435	Standard
	Cr	52	51911.1	1.7	**4.8949**	0.101	2.1	ug/L	7635	Standard
	Cr	53	5774.4	3.1	**4.9492**	0.156	3.2	ug/L	363	Standard
	Mn	55	492914.2	2.0	**59.6732**	1.219	2.0	ug/L	747	Standard
	Co	59	9193.8	6.0	**1.1391**	0.070	6.1	ug/L	234	Standard
	Ni	60	6665.5	3.0	**2.8840**	0.083	2.9	ug/L	215	Standard
	Cu	65	17774.1	0.7	**7.9654**	0.025	0.3	ug/L	104	Standard
	Zn	66	87339.7	0.8	**67.1049**	0.286	0.4	ug/L	363	Standard
>	Ge	72	585998.9	0.5				ug/L	495142	Standard
	As	75	970.2	5.1	**0.7914**	0.035	4.4	ug/L	-50	Standard
	Se	82	25.1	35.6	**0.0547**	0.068	124.2	ug/L	14	Standard
	Se-1	77	82.0	5.6	**0.3262**	0.050	15.4	ug/L	42	Standard
>	Ga	71	2373.5	5.4				mg/L	42	Standard
	Rb	85	23598.4	1.4				ug/L	17	Standard
	Y	89	515408.4	0.8				ug/L	417098	Standard
>	Rh	103	21.7	35.3				ug/L	13	Standard
	Mo	98	369.9	37.1	**0.1174**	0.046	39.0	ug/L	11	Standard
	Ag	107	699.0	38.5	**0.0822**	0.037	44.7	ug/L	68	Standard
	Cd	111	3080.1	3.0	**1.2985**	0.040	3.1	mg/L	2	Standard
	Cd	114	7389.1	5.7	**1.2756**	0.072	5.6	ug/L	12	Standard
>	In	115	662403.8	0.2				ug/L	587441	Standard
	Sn	118	660.0	13.5	**0.0055**	0.013	245.5	ug/L	497	Standard
	Sb	123	2141.2	12.8	**0.4025**	0.052	13.0	ug/L	78	Standard
	Ba	135	63716.0	0.4	**27.5914**	0.127	0.5	ug/L	24	Standard
	Ce	140	193219.3	1.3				ug/L	25	Standard
>	Tb	159	1027182.9	1.4				ug/L	939701	Standard
	Ho	165	2611.9	6.2				ug/L	10	Standard
	Tl	203	353.7	43.7	**0.0374**	0.017	46.4	ug/L	4	Standard
	Tl	205	421.7	67.1	**0.0541**	0.036	67.5	ug/L	2	Standard
	Pb	206	67035.7	2.8	**11.3788**	0.381	3.4	ug/L	240	Standard
	Pb	207	53097.6	2.6	**10.1377**	0.324	3.2	ug/L	198	Standard
	Pb	208	229986.8	1.7	**11.1165**	0.256	2.3	ug/L	883	Standard
	U	238	1351.4	18.8	**0.1748**	0.035	19.9	ug/L	3	Standard
>	Bi	209	574350.4	0.6				ug/L	523984	Standard

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:34:10
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	130.0	13.3	**0.0772**	0.020	26.0	mg/L	37	Standard
K	39	128.3	6.0	**0.1263**	0.018	14.2	mg/L	38	Standard
Ca	43	216.7	23.7	**9.0911**	2.889	31.8	mg/L	10	Standard
Fe	54	2661.6	0.9	**1.7733**	0.039	2.2	mg/L	179	Standard
Fe	57	965.0	5.3	**1.7027**	0.101	5.9	mg/L	335	Standard
> **Sc-1**	45	72059.3	1.7				mg/L	61029	Standard
Cl	35	30260.9	1.5				ug/L	20195	Standard
Kr	83	0.3	173.2				ug/L	4	Standard
Br	81	2210.2	10.2				ug/L	1547	Standard
P	31	20780.9	2.4				ug/L	15286	Standard
S	34	2430.2	2.3				ug/L	1572	Standard
Sr	88	83.3	13.9				ug/L	73	Standard
C	12	56.7	40.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	33.3	17.3				mg/L	3	Standard
Dy	164	4484.9	5.6				mg/L	6	Standard
Ho-1	165	2611.9	6.2				mg/L	10	Standard
Er	166	2556.9	13.6				mg/L	13	Standard
I	127	10483.6	6.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.090	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		118.350	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:34:10
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.761
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.612
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120605SDL WG574690-02
Sample Date/Time: Wednesday, June 29, 2016 18:35:04
Number of Replicates: 3
Autosampler Position: 246
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83239.9	3.6				ug/L	74699	Standard
	Be	9	38.3	32.8	**0.0467**	0.013	27.9	ug/L	8	Standard
	Al	27	353337.7	1.0	**2.1460**	0.078	3.6	ug/L	1493	Standard
	Sc	45	70061.7	2.7				ug/L	61029	Standard
	Ti	47	444.3	5.4	**0.9019**	0.065	7.2	ug/L	35	Standard
	V	51	7638.3	1.2	**0.6761**	0.009	1.3	ug/L	1435	Standard
	Cr	52	17382.6	1.1	**0.9981**	0.034	3.5	ug/L	7635	Standard
	Cr	53	1533.4	2.6	**1.0999**	0.009	0.8	ug/L	363	Standard
	Mn	55	100723.4	1.5	**12.3208**	0.101	0.8	ug/L	747	Standard
	Co	59	1988.5	0.8	**0.2223**	0.008	3.5	ug/L	234	Standard
	Ni	60	1493.7	0.8	**0.5638**	0.012	2.1	ug/L	215	Standard
	Cu	65	3674.4	3.3	**1.5981**	0.060	3.8	ug/L	104	Standard
	Zn	66	19238.9	1.8	**14.7255**	0.173	1.2	ug/L	363	Standard
>	**Ge**	72	574541.2	2.2				ug/L	495142	Standard
	As	75	156.2	2.1	**0.1799**	0.005	2.7	ug/L	-50	Standard
	Se	82	23.9	29.5	**0.0491**	0.053	108.8	ug/L	14	Standard
	Se-1	77	58.0	21.2	**0.0603**	0.130	215.1	ug/L	42	Standard
>	**Ga**	71	518.3	4.3				mg/L	42	Standard
	Rb	85	4807.4	2.9				ug/L	17	Standard
	Y	89	501442.7	1.9				ug/L	417098	Standard
>	**Rh**	103	25.0	20.0				ug/L	13	Standard
	Mo	98	84.6	15.3	**0.0217**	0.004	18.8	ug/L	11	Standard
	Ag	107	254.3	9.6	**0.0208**	0.003	15.4	ug/L	68	Standard
	Cd	111	620.8	2.0	**0.2487**	0.004	1.5	mg/L	2	Standard
	Cd	114	1552.7	9.2	**0.2596**	0.025	9.8	ug/L	12	Standard
>	**In**	115	679737.8	0.5				ug/L	587441	Standard
	Sn	118	525.0	13.7	**-0.0171**	0.010	60.3	ug/L	497	Standard
	Sb	123	593.1	26.9	**0.1056**	0.030	28.3	ug/L	78	Standard
	Ba	135	12802.7	1.3	**5.3901**	0.099	1.8	ug/L	24	Standard
	Ce	140	39412.9	3.5				ug/L	25	Standard
>	**Tb**	159	1032176.9	1.0				ug/L	939701	Standard
	Ho	165	550.0	19.7				ug/L	10	Standard
	Tl	203	66.7	15.4	**0.0056**	0.001	20.8	ug/L	4	Standard
	Tl	205	78.3	41.0	**0.0102**	0.004	40.5	ug/L	2	Standard
	Pb	206	13522.4	0.6	**2.2309**	0.023	1.0	ug/L	240	Standard
	Pb	207	10801.2	2.0	**1.9994**	0.022	1.1	ug/L	198	Standard
	Pb	208	47468.1	0.8	**2.2283**	0.042	1.9	ug/L	883	Standard
	U	238	258.3	8.4	**0.0297**	0.002	8.2	ug/L	3	Standard
>	**Bi**	209	576211.7	1.3				ug/L	523984	Standard

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:37:21
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	16.4	**-0.0125**	0.009	74.1	mg/L	37	Standard
K	39	58.3	35.7	**-0.0022**	0.039	1787.9	mg/L	38	Standard
Ca	43	95.0	32.0	**2.7333**	1.777	65.0	mg/L	10	Standard
Fe	54	697.3	2.7	**0.3769**	0.020	5.3	mg/L	179	Standard
Fe	57	380.0	7.9	**0.0519**	0.104	201.0	mg/L	335	Standard
> **Sc-1**	45	70061.7	2.7				mg/L	61029	Standard
Cl	35	30787.9	1.4				ug/L	20195	Standard
Kr	83	2.3	107.9				ug/L	4	Standard
Br	81	2130.1	1.6				ug/L	1547	Standard
P	31	20966.2	4.2				ug/L	15286	Standard
S	34	2485.2	3.8				ug/L	1572	Standard
Sr	88	90.0	33.3				ug/L	73	Standard
C	12	40.0	66.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	3	Standard
Dy	164	970.4	12.1				mg/L	6	Standard
Ho-1	165	550.0	19.7				mg/L	10	Standard
Er	166	480.0	14.6				mg/L	13	Standard
I	127	3620.4	13.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.434	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		116.036	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120605SDL WG574690-02
Report Date/Time: Wednesday, June 29, 2016 18:37:21
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		115.712
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.967
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 18:38:18
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	82768.9	2.2				ug/L	74699	Standard
	Be	9	55586.8	2.9	**56.9851**	0.437	0.8	ug/L	8	Standard
	Al	27	8619622.5	1.7	**52.9268**	0.666	1.3	ug/L	1493	Standard
	Sc	45	69535.9	1.6				ug/L	61029	Standard
	Ti	47	45959.0	2.5	**108.3957**	2.700	2.5	ug/L	35	Standard
	V	51	452339.2	1.9	**53.3985**	1.095	2.1	ug/L	1435	Standard
	Cr	52	452917.5	1.6	**53.6621**	0.813	1.5	ug/L	7635	Standard
	Cr	53	54972.9	2.4	**52.8966**	1.385	2.6	ug/L	363	Standard
	Mn	55	414134.0	1.2	**53.0785**	0.725	1.4	ug/L	747	Standard
	Co	59	411024.9	0.5	**55.6590**	0.257	0.5	ug/L	234	Standard
	Ni	60	111206.0	1.9	**53.0242**	0.984	1.9	ug/L	215	Standard
	Cu	65	110023.1	1.1	**52.7940**	0.716	1.4	ug/L	104	Standard
	Zn	66	64745.4	0.8	**52.5833**	0.259	0.5	ug/L	363	Standard
>	Ge	72	553341.4	0.3				ug/L	495142	Standard
	As	75	65551.1	0.4	**52.4265**	0.227	0.4	ug/L	-50	Standard
	Se	82	6664.4	1.0	**54.1963**	0.405	0.7	ug/L	14	Standard
	Se-1	77	4501.0	1.7	**54.5222**	0.959	1.8	ug/L	42	Standard
>	Ga	71	63.3	38.9				mg/L	42	Standard
	Rb	85	948.4	4.5				ug/L	17	Standard
	Y	89	490323.1	1.2				ug/L	417098	Standard
>	Rh	103	18.3	41.7				ug/L	13	Standard
	Mo	98	304786.5	0.3	**99.6729**	0.315	0.3	ug/L	11	Standard
	Ag	107	371315.6	1.3	**49.6520**	0.733	1.5	ug/L	68	Standard
	Cd	111	119443.2	1.0	**49.7891**	0.404	0.8	mg/L	2	Standard
	Cd	114	290803.0	0.5	**49.4253**	0.213	0.4	ug/L	12	Standard
>	In	115	673907.8	0.2				ug/L	587441	Standard
	Sn	118	339134.4	2.1	**50.5123**	1.079	2.1	ug/L	497	Standard
	Sb	123	262026.6	0.8	**48.9184**	0.455	0.9	ug/L	78	Standard
	Ba	135	111822.5	0.7	**47.6088**	0.434	0.9	ug/L	24	Standard
	Ce	140	206.7	14.6				ug/L	25	Standard
>	Tb	159	1024624.4	0.5				ug/L	939701	Standard
	Ho	165	8.3	34.6				ug/L	10	Standard
	Tl	203	457896.5	0.3	**51.5737**	0.953	1.8	ug/L	4	Standard
	Tl	205	420019.2	1.0	**54.5041**	0.795	1.5	ug/L	2	Standard
	Pb	206	290158.6	1.2	**50.3552**	0.541	1.1	ug/L	240	Standard
	Pb	207	256755.7	1.2	**50.1530**	0.630	1.3	ug/L	198	Standard
	Pb	208	1071548.3	0.3	**52.9883**	1.261	2.4	ug/L	883	Standard
	U	238	389229.1	0.1	**52.5000**	1.135	2.2	ug/L	3	Standard
>	Bi	209	564495.9	2.1				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 18:40:35
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.7614**	29.023	173.2	mg/L	0	Standard
Mg	24	4642.4	0.2	**5.1955**	0.100	1.9	mg/L	37	Standard
K	39	2168.5	2.0	**4.0939**	0.144	3.5	mg/L	38	Standard
Ca	43	86.7	35.3	**2.2659**	1.624	71.7	mg/L	10	Standard
Fe	54	7113.4	2.2	**5.1545**	0.030	0.6	mg/L	179	Standard
Fe	57	2041.8	2.3	**5.0193**	0.241	4.8	mg/L	335	Standard
> **Sc-1**	45	69535.9	1.6				mg/L	61029	Standard
Cl	35	30041.8	1.0				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	1843.4	4.1				ug/L	1547	Standard
P	31	21683.8	2.5				ug/L	15286	Standard
S	34	2823.6	5.0				ug/L	1572	Standard
Sr	88	53.3	39.0				ug/L	73	Standard
C	12	53.3	78.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	12.9	179.7				mg/L	6	Standard
Ho-1	165	8.3	34.6				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	2093.5	24.6				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	113.970		
	Al	27	105.854		
	Sc	45			
	Ti	47	108.396		
	V	51	106.797		
	Cr	52	107.324		
	Cr	53			
	Mn	55	106.157		
	Co	59	111.318		
	Ni	60	106.048		
	Cu	65	105.588		
	Zn	66	105.167		
>	Ge	72		111.754	
	As	75	104.853		
	Se	82	108.393		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 18:40:35
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	99.673		
Ag	107	99.304		
Cd	111	99.578		
Cd	114			
In	115		114.719	
Sn	118	101.025		
Sb	123	97.837		
Ba	135	95.218		
Ce	140			
Tb	159			
Ho	165			
Tl	203	103.147		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.977		
U	238	105.000		
Bi	209		107.731	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 18:40:35
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 18:41:29
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83048.8	2.8				ug/L	74699	Standard
	Be	9	23.3	68.9	**0.0313**	0.016	52.2	ug/L	8	Standard
	Al	27	3078.7	25.1	**0.0054**	0.004	78.6	ug/L	1493	Standard
	Sc	45	71367.7	1.6				ug/L	61029	Standard
	Ti	47	35.3	33.5	**-0.0277**	0.025	91.3	ug/L	35	Standard
	V	51	1419.9	5.2	**-0.0309**	0.006	18.5	ug/L	1435	Standard
	Cr	52	7844.0	0.2	**-0.0958**	0.021	21.7	ug/L	7635	Standard
	Cr	53	443.3	5.7	**0.0910**	0.034	37.3	ug/L	363	Standard
	Mn	55	990.7	12.0	**-0.0236**	0.012	50.6	ug/L	747	Standard
	Co	59	281.0	9.0	**-0.0000**	0.003	39793.0	ug/L	234	Standard
	Ni	60	196.7	7.3	**-0.0319**	0.005	14.4	ug/L	215	Standard
	Cu	65	145.3	12.4	**-0.0353**	0.007	21.2	ug/L	104	Standard
	Zn	66	464.0	9.4	**-0.0826**	0.026	32.0	ug/L	363	Standard
>	**Ge**	72	564975.6	2.4				ug/L	495142	Standard
	As	75	-39.3	63.0	**0.0286**	0.020	68.3	ug/L	-50	Standard
	Se	82	26.6	20.4	**0.0737**	0.038	51.8	ug/L	14	Standard
	Se-1	77	58.7	15.8	**0.0805**	0.100	124.7	ug/L	42	Standard
>	**Ga**	71	41.7	25.0				mg/L	42	Standard
	Rb	85	35.0	75.6				ug/L	17	Standard
	Y	89	486175.7	0.9				ug/L	417098	Standard
>	**Rh**	103	21.7	58.1				ug/L	13	Standard
	Mo	98	275.2	23.2	**0.0821**	0.020	24.6	ug/L	11	Standard
	Ag	107	144.3	16.7	**0.0059**	0.003	51.4	ug/L	68	Standard
	Cd	111	18.2	29.8	**-0.0005**	0.002	416.8	mg/L	2	Standard
	Cd	114	40.0	61.5	**0.0045**	0.004	89.7	ug/L	12	Standard
>	**In**	115	690622.1	0.6				ug/L	587441	Standard
	Sn	118	806.7	10.4	**0.0228**	0.013	56.6	ug/L	497	Standard
	Sb	123	553.4	29.9	**0.0967**	0.030	31.3	ug/L	78	Standard
	Ba	135	70.0	42.9	**0.0130**	0.012	94.6	ug/L	24	Standard
	Ce	140	61.7	54.0				ug/L	25	Standard
>	**Tb**	159	1047202.6	1.6				ug/L	939701	Standard
	Ho	165	16.7	34.6				ug/L	10	Standard
	Tl	203	75.3	48.5	**0.0063**	0.004	58.9	ug/L	4	Standard
	Tl	205	73.3	27.6	**0.0093**	0.002	26.3	ug/L	2	Standard
	Pb	206	333.0	14.3	**-0.0156**	0.007	42.7	ug/L	240	Standard
	Pb	207	293.0	14.2	**-0.0150**	0.006	43.1	ug/L	198	Standard
	Pb	208	1234.7	14.5	**-0.0154**	0.007	47.2	ug/L	883	Standard
	U	238	61.3	33.0	**0.0035**	0.002	69.2	ug/L	3	Standard
>	**Bi**	209	587080.9	2.4				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:43:46
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	38.3	15.1	**-0.0227**	0.007	30.4	mg/L	37	Standard
K	39	33.3	22.9	**-0.0512**	0.014	27.0	mg/L	38	Standard
Ca	43	28.3	50.9	**-0.9868**	0.802	81.3	mg/L	10	Standard
Fe	54	132.7	11.9	**-0.0421**	0.011	25.9	mg/L	179	Standard
Fe	57	280.0	10.7	**-0.2593**	0.100	38.4	mg/L	335	Standard
Sc-1	45	71367.7	1.6				mg/L	61029	Standard
Cl	35	29768.6	1.4				ug/L	20195	Standard
Kr	83	1.3	173.2				ug/L	4	Standard
Br	81	2106.8	7.7				ug/L	1547	Standard
P	31	21775.6	2.2				ug/L	15286	Standard
S	34	2523.5	3.0				ug/L	1572	Standard
Sr	88	70.0	49.5				ug/L	73	Standard
C	12	66.7	48.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	19.2	48.3				mg/L	6	Standard
Ho-1	165	16.7	34.6				mg/L	10	Standard
Er	166	16.7	91.7				mg/L	13	Standard
I	127	1905.1	33.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.104	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:43:46
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.564
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		112.042
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 18:43:46
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120701
Sample Date/Time: Wednesday, June 29, 2016 18:44:42
Number of Replicates: 3
Autosampler Position: 247
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82529.2	0.7				ug/L	74699	Standard
	Be	9	748.4	11.5	**0.7773**	0.092	11.9	ug/L	8	Standard
	Al	27	5186124.5	1.1	**31.9282**	0.368	1.2	ug/L	1493	Standard
	Sc	45	73214.8	0.3				ug/L	61029	Standard
	Ti	47	12465.8	1.5	**28.7779**	0.271	0.9	ug/L	35	Standard
	V	51	254152.3	1.7	**29.3595**	0.344	1.2	ug/L	1435	Standard
	Cr	52	145419.2	1.0	**16.2106**	0.127	0.8	ug/L	7635	Standard
	Cr	53	17483.4	2.8	**16.2818**	0.208	1.3	ug/L	363	Standard
	Mn	55	3993207.5	1.0	**503.6001**	3.344	0.7	ug/L	747	Standard
	Co	59	68748.0	1.0	**9.1067**	0.083	0.9	ug/L	234	Standard
	Ni	60	27938.1	3.3	**12.9787**	0.214	1.6	ug/L	215	Standard
	Cu	65	24664.7	1.5	**11.5352**	0.050	0.4	ug/L	104	Standard
	Zn	66	148101.8	1.7	**118.6204**	0.288	0.2	ug/L	363	Standard
>	**Ge**	72	563786.2	1.7				ug/L	495142	Standard
	As	75	6453.5	1.8	**5.1196**	0.031	0.6	ug/L	-50	Standard
	Se	82	62.5	2.1	**0.3624**	0.012	3.2	ug/L	14	Standard
	Se-1	77	114.0	16.7	**0.7504**	0.247	32.9	ug/L	42	Standard
>	**Ga**	71	11409.3	2.4				mg/L	42	Standard
	Rb	85	132848.2	0.6				ug/L	17	Standard
	Y	89	719521.7	0.4				ug/L	417098	Standard
>	**Rh**	103	43.3	43.7				ug/L	13	Standard
	Mo	98	809.7	3.4	**0.2640**	0.011	4.2	ug/L	11	Standard
	Ag	107	5298.9	1.5	**0.7088**	0.004	0.5	ug/L	68	Standard
	Cd	111	3980.4	2.6	**1.6822**	0.055	3.3	mg/L	2	Standard
	Cd	114	9713.7	1.4	**1.6793**	0.030	1.8	ug/L	12	Standard
>	**In**	115	661793.4	1.1				ug/L	587441	Standard
	Sn	118	911.7	2.8	**0.0439**	0.005	10.9	ug/L	497	Standard
	Sb	123	350.1	17.1	**0.0624**	0.012	19.0	ug/L	78	Standard
	Ba	135	341983.5	0.6	**148.3035**	0.748	0.5	ug/L	24	Standard
	Ce	140	2710163.1	0.4				ug/L	25	Standard
>	**Tb**	159	1034801.7	0.1				ug/L	939701	Standard
	Ho	165	19319.0	1.7				ug/L	10	Standard
	Tl	203	872.7	2.5	**0.0967**	0.002	2.5	ug/L	4	Standard
	Tl	205	843.4	5.5	**0.1098**	0.006	5.6	ug/L	2	Standard
	Pb	206	283131.4	1.0	**49.2257**	0.472	1.0	ug/L	240	Standard
	Pb	207	230315.5	1.3	**45.0642**	0.519	1.2	ug/L	198	Standard
	Pb	208	999318.0	0.9	**49.4947**	0.507	1.0	ug/L	883	Standard
	U	238	4899.8	1.7	**0.6576**	0.010	1.6	ug/L	3	Standard
>	**Bi**	209	563366.5	0.2				ug/L	523984	Standard

Sample ID: L1606120701
Report Date/Time: Wednesday, June 29, 2016 18:46:59
Page 1

Approved: July 04, 2016

Na	23	5.0	0.0	**48.0405**	0.159	0.3	mg/L	0	Standard
Mg	24	875.0	7.7	**0.8766**	0.075	8.6	mg/L	37	Standard
K	39	536.7	5.1	**0.8747**	0.054	6.2	mg/L	38	Standard
Ca	43	571.7	16.7	**27.5670**	4.925	17.9	mg/L	10	Standard
Fe	54	14788.8	1.9	**10.3138**	0.228	2.2	mg/L	179	Standard
Fe	57	4073.9	8.9	**10.4697**	1.059	10.1	mg/L	335	Standard
> **Sc-1**	45	73214.8	0.3				mg/L	61029	Standard
Cl	35	28264.3	0.5				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	3023.6	7.8				ug/L	1547	Standard
P	31	20599.0	2.2				ug/L	15286	Standard
S	34	2595.2	1.7				ug/L	1572	Standard
Sr	88	203.3	16.4				ug/L	73	Standard
C	12	130.0	40.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	150.0	43.7				mg/L	3	Standard
Dy	164	31532.4	1.2				mg/L	6	Standard
Ho-1	165	19319.0	1.7				mg/L	10	Standard
Er	166	17720.4	2.1				mg/L	13	Standard
I	127	25842.0	0.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.483	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.863	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120701
Report Date/Time: Wednesday, June 29, 2016 18:46:59
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	112.657
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	107.516
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120701
Report Date/Time: Wednesday, June 29, 2016 18:46:59
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120702
Sample Date/Time: Wednesday, June 29, 2016 18:47:54
Number of Replicates: 3
Autosampler Position: 248
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85381.8	3.5				ug/L	74699	Standard
	Be	9	623.3	3.3	**0.6272**	0.018	2.8	ug/L	8	Standard
	Al	27	4620203.8	1.9	**27.5041**	0.647	2.4	ug/L	1493	Standard
	Sc	45	75576.5	3.2				ug/L	61029	Standard
	Ti	47	6206.6	0.5	**14.3105**	0.192	1.3	ug/L	35	Standard
	V	51	134133.3	1.8	**15.4410**	0.153	1.0	ug/L	1435	Standard
	Cr	52	96500.3	0.7	**10.4416**	0.108	1.0	ug/L	7635	Standard
	Cr	53	11224.2	2.7	**10.3669**	0.448	4.3	ug/L	363	Standard
	Mn	55	3577332.4	1.7	**452.2740**	8.609	1.9	ug/L	747	Standard
	Co	59	47345.9	1.4	**6.2757**	0.086	1.4	ug/L	234	Standard
	Ni	60	24689.4	1.0	**11.4869**	0.101	0.9	ug/L	215	Standard
	Cu	65	29948.9	2.1	**14.0640**	0.244	1.7	ug/L	104	Standard
	Zn	66	164514.2	1.7	**132.1464**	1.033	0.8	ug/L	363	Standard
>	**Ge**	72	562390.3	1.6				ug/L	495142	Standard
	As	75	2776.0	3.7	**2.2412**	0.056	2.5	ug/L	-50	Standard
	Se	82	106.3	15.9	**0.7139**	0.124	17.3	ug/L	14	Standard
	Se-1	77	132.3	11.0	**0.9743**	0.190	19.5	ug/L	42	Standard
>	**Ga**	71	12940.5	1.1				mg/L	42	Standard
	Rb	85	250383.8	1.2				ug/L	17	Standard
	Y	89	695657.2	0.8				ug/L	417098	Standard
>	**Rh**	103	35.0	62.3				ug/L	13	Standard
	Mo	98	1645.8	3.3	**0.5188**	0.023	4.4	ug/L	11	Standard
	Ag	107	671.7	7.7	**0.0746**	0.007	9.4	ug/L	68	Standard
	Cd	111	2017.5	2.8	**0.8116**	0.028	3.5	mg/L	2	Standard
	Cd	114	4822.6	1.0	**0.7964**	0.005	0.7	ug/L	12	Standard
>	**In**	115	691832.3	1.1				ug/L	587441	Standard
	Sn	118	616.7	6.1	**-0.0050**	0.006	126.3	ug/L	497	Standard
	Sb	123	493.2	8.1	**0.0855**	0.008	9.1	ug/L	78	Standard
	Ba	135	281871.8	0.7	**116.9321**	1.732	1.5	ug/L	24	Standard
	Ce	140	1017603.0	1.2				ug/L	25	Standard
>	**Tb**	159	1062512.9	1.9				ug/L	939701	Standard
	Ho	165	14416.9	1.8				ug/L	10	Standard
	Tl	203	1059.0	3.0	**0.1166**	0.005	4.1	ug/L	4	Standard
	Tl	205	896.7	14.9	**0.1158**	0.019	16.8	ug/L	2	Standard
	Pb	206	100858.6	1.2	**17.3191**	0.532	3.1	ug/L	240	Standard
	Pb	207	81475.8	1.0	**15.7413**	0.458	2.9	ug/L	198	Standard
	Pb	208	356757.6	0.9	**17.4487**	0.446	2.6	ug/L	883	Standard
	U	238	6683.8	3.5	**0.8902**	0.048	5.4	ug/L	3	Standard
>	**Bi**	209	569106.4	1.9				ug/L	523984	Standard

Sample ID: L1606120702
Report Date/Time: Wednesday, June 29, 2016 18:50:11
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**31.5301**	27.313	86.6	mg/L	0	Standard
Mg	24	745.0	7.0	**0.7121**	0.061	8.5	mg/L	37	Standard
K	39	773.4	6.0	**1.2683**	0.120	9.5	mg/L	38	Standard
Ca	43	171.7	3.4	**6.2441**	0.546	8.7	mg/L	10	Standard
Fe	54	10724.1	1.5	**7.2104**	0.318	4.4	mg/L	179	Standard
Fe	57	2831.9	1.0	**6.7023**	0.187	2.8	mg/L	335	Standard
> **Sc-1**	45	75576.5	3.2				mg/L	61029	Standard
Cl	35	27596.4	1.5				ug/L	20195	Standard
Kr	83	4.0	66.1				ug/L	4	Standard
Br	81	8432.3	2.7				ug/L	1547	Standard
P	31	20674.1	3.5				ug/L	15286	Standard
S	34	2360.2	7.8				ug/L	1572	Standard
Sr	88	120.0	18.2				ug/L	73	Standard
C	12	350.0	15.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	106.7	43.3				mg/L	3	Standard
Dy	164	23142.4	1.5				mg/L	6	Standard
Ho-1	165	14416.9	1.8				mg/L	10	Standard
Er	166	13422.6	4.3				mg/L	13	Standard
I	127	174464.7	3.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		114.301	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.582	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120702
Report Date/Time: Wednesday, June 29, 2016 18:50:11
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	117.770
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.611
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120702
Report Date/Time: Wednesday, June 29, 2016 18:50:11
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120801
Sample Date/Time: Wednesday, June 29, 2016 18:51:05
Number of Replicates: 3
Autosampler Position: 249
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83030.4	3.3				ug/L	74699	Standard
	Be	9	1115.0	11.0	**1.1453**	0.091	8.0	ug/L	8	Standard
	Al	27	17314833.4	0.9	**106.0421**	2.666	2.5	ug/L	1493	Standard
	Sc	45	77183.1	5.1				ug/L	61029	Standard
	Ti	47	40548.2	1.5	**95.7436**	2.793	2.9	ug/L	35	Standard
	V	51	251782.0	1.4	**29.6946**	1.634	5.5	ug/L	1435	Standard
	Cr	52	243737.7	2.0	**28.4388**	0.710	2.5	ug/L	7635	Standard
	Cr	53	29872.8	2.3	**28.6239**	0.690	2.4	ug/L	363	Standard
	Mn	55	3685844.9	1.5	**474.5740**	27.736	5.8	ug/L	747	Standard
	Co	59	66271.6	1.6	**8.9620**	0.528	5.9	ug/L	234	Standard
	Ni	60	46026.2	2.6	**21.8948**	0.433	2.0	ug/L	215	Standard
	Cu	65	42002.8	1.5	**20.1164**	0.608	3.0	ug/L	104	Standard
	Zn	66	111824.0	2.6	**91.2379**	1.710	1.9	ug/L	363	Standard
>	**Ge**	72	553103.7	4.4				ug/L	495142	Standard
	As	75	5557.5	3.1	**4.5033**	0.093	2.1	ug/L	-50	Standard
	Se	82	102.2	5.1	**0.6977**	0.072	10.3	ug/L	14	Standard
	Se-1	77	134.3	12.1	**1.0314**	0.270	26.2	ug/L	42	Standard
>	**Ga**	71	24266.1	1.0				mg/L	42	Standard
	Rb	85	181545.0	2.7				ug/L	17	Standard
	Y	89	698995.8	3.0				ug/L	417098	Standard
>	**Rh**	103	88.3	6.5				ug/L	13	Standard
	Mo	98	2397.1	2.7	**0.8492**	0.049	5.8	ug/L	11	Standard
	Ag	107	1012.7	3.0	**0.1348**	0.009	6.5	ug/L	68	Standard
	Cd	111	911.2	3.1	**0.4060**	0.012	2.9	mg/L	2	Standard
	Cd	114	2080.2	4.3	**0.3830**	0.011	2.9	ug/L	12	Standard
>	**In**	115	618664.5	3.1				ug/L	587441	Standard
	Sn	118	1240.1	3.1	**0.1071**	0.013	12.0	ug/L	497	Standard
	Sb	123	1640.7	2.7	**0.3299**	0.019	5.8	ug/L	78	Standard
	Ba	135	534847.9	1.0	**248.3132**	9.882	4.0	ug/L	24	Standard
	Ce	140	1227932.9	1.4				ug/L	25	Standard
>	**Tb**	159	991058.9	3.3				ug/L	939701	Standard
	Ho	165	19152.1	0.5				ug/L	10	Standard
	Tl	203	1084.7	2.8	**0.1294**	0.004	2.8	ug/L	4	Standard
	Tl	205	976.7	6.4	**0.1361**	0.004	2.8	ug/L	2	Standard
	Pb	206	122644.8	1.6	**22.8364**	1.340	5.9	ug/L	240	Standard
	Pb	207	98294.5	1.0	**20.5914**	1.111	5.4	ug/L	198	Standard
	Pb	208	423360.3	0.4	**22.4444**	1.000	4.5	ug/L	883	Standard
	U	238	7556.2	1.8	**1.0902**	0.045	4.1	ug/L	3	Standard
>	**Bi**	209	526041.8	4.3				ug/L	523984	Standard

Sample ID: L1606120801
Report Date/Time: Wednesday, June 29, 2016 18:53:22
Page 1

Approved: July 04, 2016

Na	23	10.0	50.0	**92.8331**	50.540	54.4	mg/L	0	Standard
Mg	24	373.3	13.0	**0.3164**	0.050	15.9	mg/L	37	Standard
K	39	846.7	3.0	**1.3673**	0.064	4.7	mg/L	38	Standard
Ca	43	1881.8	8.7	**91.3781**	4.657	5.1	mg/L	10	Standard
Fe	54	16605.4	3.8	**11.0236**	0.942	8.5	mg/L	179	Standard
Fe	57	5247.6	6.9	**13.0290**	0.645	4.9	mg/L	335	Standard
Sc-1	45	77183.1	5.1				mg/L	61029	Standard
Cl	35	28165.5	0.9				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	3540.4	14.7				ug/L	1547	Standard
P	31	23239.5	5.6				ug/L	15286	Standard
S	34	2426.9	3.5				ug/L	1572	Standard
Sr	88	415.0	11.5				ug/L	73	Standard
C	12	173.3	6.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	66.7	60.6				mg/L	3	Standard
Dy	164	30988.1	3.0				mg/L	6	Standard
Ho-1	165	19152.1	0.5				mg/L	10	Standard
Er	166	17717.0	2.5				mg/L	13	Standard
I	127	27493.3	1.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		111.154	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.706	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606120801
Report Date/Time: Wednesday, June 29, 2016 18:53:22
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	105.315
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	100.393
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120801
Report Date/Time: Wednesday, June 29, 2016 18:53:22
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606120802
Sample Date/Time: Wednesday, June 29, 2016 18:54:17
Number of Replicates: 3
Autosampler Position: 250
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88923.6	2.8				ug/L	74699	Standard
	Be	9	661.7	11.4	**0.6387**	0.067	10.5	ug/L	8	Standard
	Al	27	15492795.6	1.0	**88.5762**	1.812	2.0	ug/L	1493	Standard
	Sc	45	80386.0	1.0				ug/L	61029	Standard
	Ti	47	20524.2	1.6	**45.7756**	1.470	3.2	ug/L	35	Standard
	V	51	220306.5	1.3	**24.5176**	0.534	2.2	ug/L	1435	Standard
	Cr	52	643370.5	1.6	**72.5457**	2.733	3.8	ug/L	7635	Standard
	Cr	53	78035.6	3.0	**71.2237**	3.200	4.5	ug/L	363	Standard
	Mn	55	3332205.3	1.0	**405.3306**	7.931	2.0	ug/L	747	Standard
	Co	59	49008.5	0.3	**6.2510**	0.159	2.5	ug/L	234	Standard
	Ni	60	36936.6	2.2	**16.5952**	0.689	4.2	ug/L	215	Standard
	Cu	65	48506.2	1.8	**21.9786**	0.680	3.1	ug/L	104	Standard
	Zn	66	431997.7	0.4	**334.6356**	8.273	2.5	ug/L	363	Standard
>	**Ge**	72	584593.6	2.3				ug/L	495142	Standard
	As	75	4706.8	0.7	**3.6196**	0.063	1.8	ug/L	-50	Standard
	Se	82	170.0	3.4	**1.1737**	0.017	1.5	ug/L	14	Standard
	Se-1	77	180.7	11.7	**1.4743**	0.259	17.6	ug/L	42	Standard
>	**Ga**	71	19881.4	1.3				mg/L	42	Standard
	Rb	85	200103.0	1.8				ug/L	17	Standard
	Y	89	657665.4	0.6				ug/L	417098	Standard
>	**Rh**	103	171.7	7.3				ug/L	13	Standard
	Mo	98	1216.9	1.5	**0.3934**	0.011	2.8	ug/L	11	Standard
	Ag	107	1006.0	3.7	**0.1221**	0.007	5.4	ug/L	68	Standard
	Cd	111	4519.9	2.1	**1.8818**	0.049	2.6	mg/L	2	Standard
	Cd	114	11256.2	1.6	**1.9163**	0.030	1.6	ug/L	12	Standard
>	**In**	115	672127.4	1.5				ug/L	587441	Standard
	Sn	118	1105.0	15.3	**0.0707**	0.026	36.7	ug/L	497	Standard
	Sb	123	229.1	10.3	**0.0387**	0.005	12.5	ug/L	78	Standard
	Ba	135	361495.2	1.1	**154.3886**	3.928	2.5	ug/L	24	Standard
	Ce	140	936346.8	0.9				ug/L	25	Standard
>	**Tb**	159	1070983.4	3.2				ug/L	939701	Standard
	Ho	165	13711.2	2.1				ug/L	10	Standard
	Tl	203	1023.7	3.8	**0.1103**	0.006	5.8	ug/L	4	Standard
	Tl	205	995.0	4.8	**0.1256**	0.007	5.2	ug/L	2	Standard
	Pb	206	1451242.4	0.5	**244.9779**	5.608	2.3	ug/L	240	Standard
	Pb	207	1189055.7	0.7	**225.9168**	5.756	2.5	ug/L	198	Standard
	Pb	208	5216722.9	1.2	**250.8847**	7.037	2.8	ug/L	883	Standard
	U	238	9013.7	1.2	**1.1768**	0.034	2.9	ug/L	3	Standard
>	**Bi**	209	581076.6	1.9				ug/L	523984	Standard

Sample ID: L1606120802
Report Date/Time: Wednesday, June 29, 2016 18:56:34
Page 1

Approved: July 04, 2016

Na	23	6.7	173.2	**57.8631**	100.213	173.2	mg/L	0	Standard
Mg	24	553.3	14.7	**0.4776**	0.083	17.4	mg/L	37	Standard
K	39	1005.0	9.9	**1.5724**	0.167	10.6	mg/L	38	Standard
Ca	43	1240.1	6.3	**56.9237**	3.160	5.6	mg/L	10	Standard
Fe	54	17498.7	0.5	**11.1262**	0.159	1.4	mg/L	179	Standard
Fe	57	5124.2	2.5	**12.1481**	0.285	2.3	mg/L	335	Standard
Sc-1	45	80386.0	1.0				mg/L	61029	Standard
Cl	35	28790.7	0.3				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	8875.9	6.3				ug/L	1547	Standard
P	31	20292.0	8.4				ug/L	15286	Standard
S	34	2478.5	4.1				ug/L	1572	Standard
Sr	88	438.3	2.4				ug/L	73	Standard
C	12	323.3	22.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	136.7	34.6				mg/L	3	Standard
Dy	164	22618.6	1.5				mg/L	6	Standard
Ho-1	165	13711.2	2.1				mg/L	10	Standard
Er	166	12578.6	1.5				mg/L	13	Standard
I	127	73652.8	8.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		119.043	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		118.066	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606120802
Report Date/Time: Wednesday, June 29, 2016 18:56:34
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		114.416
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.896
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606120802
Report Date/Time: Wednesday, June 29, 2016 18:56:34
Page 3

Approved: July 04, 2016

Pb 206 Upper, S, EEE	Pb	206
Pb 207 Upper, S, EEE	Pb	207
Pb 208 Upper, S, EEE	Pb	208

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606129101
Sample Date/Time: Wednesday, June 29, 2016 18:57:28
Number of Replicates: 3
Autosampler Position: 251
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89580.8	1.7				ug/L	74699	Standard
	Be	9	360.0	14.7	**0.3489**	0.053	15.3	ug/L	8	Standard
	Al	27	3187676.8	1.0	**18.0772**	0.341	1.9	ug/L	1493	Standard
	Sc	45	73445.9	1.6				ug/L	61029	Standard
	Ti	47	23807.7	6.7	**54.3513**	2.712	5.0	ug/L	35	Standard
	V	51	62405.7	2.9	**6.9715**	0.072	1.0	ug/L	1435	Standard
	Cr	52	100378.5	1.9	**10.7261**	0.052	0.5	ug/L	7635	Standard
	Cr	53	11571.1	3.7	**10.5314**	0.357	3.4	ug/L	363	Standard
	Mn	55	3995310.3	0.7	**497.7426**	6.886	1.4	ug/L	747	Standard
	Co	59	12689.3	1.5	**1.6298**	0.018	1.1	ug/L	234	Standard
	Ni	60	15702.5	1.8	**7.1516**	0.026	0.4	ug/L	215	Standard
	Cu	65	1030601.6	2.5	**480.2115**	2.610	0.5	ug/L	104	Standard
	Zn	66	647092.9	2.6	**513.3926**	3.363	0.7	ug/L	363	Standard
>	**Ge**	72	570780.0	2.0				ug/L	495142	Standard
	As	75	23305.2	1.3	**18.1104**	0.138	0.8	ug/L	-50	Standard
	Se	82	481.4	2.5	**3.6678**	0.121	3.3	ug/L	14	Standard
	Se-1	77	377.3	9.1	**3.8600**	0.402	10.4	ug/L	42	Standard
>	**Ga**	71	43582.7	4.5				mg/L	42	Standard
	Rb	85	26201.0	3.0				ug/L	17	Standard
	Y	89	551000.5	1.0				ug/L	417098	Standard
>	**Rh**	103	178.3	18.7				ug/L	13	Standard
	Mo	98	21134.4	1.3	**6.6476**	0.191	2.9	ug/L	11	Standard
	Ag	107	15204.6	2.3	**1.9441**	0.032	1.6	ug/L	68	Standard
	Cd	111	1466.1	5.3	**0.5801**	0.025	4.4	mg/L	2	Standard
	Cd	114	3712.7	3.5	**0.6050**	0.007	1.1	ug/L	12	Standard
>	**In**	115	700391.9	2.5				ug/L	587441	Standard
	Sn	118	18187.6	2.5	**2.5189**	0.131	5.2	ug/L	497	Standard
	Sb	123	7709.0	2.8	**1.3807**	0.022	1.6	ug/L	78	Standard
	Ba	135	971846.2	1.7	**398.2918**	4.112	1.0	ug/L	24	Standard
	Ce	140	138638.2	1.7				ug/L	25	Standard
>	**Tb**	159	1070571.9	3.6				ug/L	939701	Standard
	Ho	165	2720.2	8.8				ug/L	10	Standard
	Tl	203	657.3	30.6	**0.0422**	0.014	32.0	ug/L	4	Standard
	Tl	205	563.3	26.0	**0.0436**	0.011	26.1	ug/L	2	Standard
	Pb	206	89328.6	2.2	**9.1434**	0.147	1.6	ug/L	240	Standard
	Pb	207	72922.4	1.4	**8.3969**	0.107	1.3	ug/L	198	Standard
	Pb	208	316979.9	0.6	**9.2432**	0.217	2.4	ug/L	883	Standard
	U	238	78871.8	1.7	**6.3120**	0.222	3.5	ug/L	3	Standard
>	**Bi**	209	950965.0	2.0				ug/L	523984	Standard

Sample ID: L1606129101
Report Date/Time: Wednesday, June 29, 2016 18:59:45
Page 1

Approved: July 04, 2016

Na	23	5.0	173.2	**48.7719**	84.467	173.2	mg/L	0	Standard
Mg	24	613.3	9.8	**0.5934**	0.073	12.3	mg/L	37	Standard
K	39	323.3	7.3	**0.4798**	0.044	9.2	mg/L	38	Standard
Ca	43	198.3	14.8	**7.8949**	1.614	20.4	mg/L	10	Standard
Fe	54	8830.8	5.5	**6.0856**	0.392	6.4	mg/L	179	Standard
Fe	57	2530.2	2.9	**6.0730**	0.228	3.8	mg/L	335	Standard
Sc-1	45	73445.9	1.6				mg/L	61029	Standard
Cl	35	25952.9	0.3				ug/L	20195	Standard
Kr	83	2.0	100.0				ug/L	4	Standard
Br	81	8005.5	5.5				ug/L	1547	Standard
P	31	19060.3	1.8				ug/L	15286	Standard
S	34	1756.8	1.0				ug/L	1572	Standard
Sr	88	518.3	2.8				ug/L	73	Standard
C	12	1226.7	9.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	1270.1	11.0				mg/L	3	Standard
Dy	164	4157.3	8.4				mg/L	6	Standard
Ho-1	165	2720.2	8.8				mg/L	10	Standard
Er	166	2643.6	8.4				mg/L	13	Standard
I	127	191520.8	4.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		119.923	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.276	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606129101
Report Date/Time: Wednesday, June 29, 2016 18:59:45
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		119.228
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		181.487
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Cu 65 Upper, S, EEE	Cu	65	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606129101
Report Date/Time: Wednesday, June 29, 2016 18:59:45
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135	
Bi 209 Int Std for sample	Bi	209	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138101
Sample Date/Time: Wednesday, June 29, 2016 19:00:40
Number of Replicates: 3
Autosampler Position: 252
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96098.8	1.2				ug/L	74699	Standard
	Be	9	1650.1	8.9	**1.4642**	0.124	8.5	ug/L	8	Standard
	Al	27	23502208.7	0.7	**124.3024**	1.056	0.8	ug/L	1493	Standard
	Sc	45	81510.4	1.4				ug/L	61029	Standard
	Ti	47	9357.2	0.5	**21.1737**	0.376	1.8	ug/L	35	Standard
	V	51	257600.4	1.3	**29.2098**	0.781	2.7	ug/L	1435	Standard
	Cr	52	224787.2	0.6	**25.1231**	0.400	1.6	ug/L	7635	Standard
	Cr	53	27600.1	3.0	**25.4059**	0.410	1.6	ug/L	363	Standard
	Mn	55	5543699.6	0.4	**686.2289**	8.924	1.3	ug/L	747	Standard
	Co	59	117640.6	0.9	**15.3216**	0.399	2.6	ug/L	234	Standard
	Ni	60	64249.3	0.8	**29.4562**	0.340	1.2	ug/L	215	Standard
	Cu	65	57365.7	1.4	**26.4629**	0.232	0.9	ug/L	104	Standard
	Zn	66	103406.0	1.2	**81.1457**	1.055	1.3	ug/L	363	Standard
>	**Ge**	72	574477.7	1.7				ug/L	495142	Standard
	As	75	12246.4	0.5	**9.4841**	0.110	1.2	ug/L	-50	Standard
	Se	82	125.1	3.5	**0.8446**	0.024	2.8	ug/L	14	Standard
	Se-1	77	164.3	2.5	**1.3176**	0.070	5.3	ug/L	42	Standard
>	**Ga**	71	29587.2	2.2				mg/L	42	Standard
	Rb	85	313937.1	2.6				ug/L	17	Standard
	Y	89	822636.0	0.8				ug/L	417098	Standard
>	**Rh**	103	38.3	49.4				ug/L	13	Standard
	Mo	98	757.2	4.8	**0.2454**	0.010	4.2	ug/L	11	Standard
	Ag	107	953.4	1.4	**0.1164**	0.003	2.3	ug/L	68	Standard
	Cd	111	1588.5	0.6	**0.6638**	0.009	1.4	mg/L	2	Standard
	Cd	114	3783.2	0.8	**0.6501**	0.009	1.4	ug/L	12	Standard
>	**In**	115	664448.0	1.0				ug/L	587441	Standard
	Sn	118	643.3	5.2	**0.0026**	0.004	152.2	ug/L	497	Standard
	Sb	123	269.8	13.5	**0.0469**	0.007	15.6	ug/L	78	Standard
	Ba	135	662128.9	0.5	**286.0118**	2.987	1.0	ug/L	24	Standard
	Ce	140	1687892.1	1.1				ug/L	25	Standard
>	**Tb**	159	1062435.8	0.8				ug/L	939701	Standard
	Ho	165	27182.7	2.2				ug/L	10	Standard
	Tl	203	1364.1	4.3	**0.1521**	0.007	4.3	ug/L	4	Standard
	Tl	205	1335.1	6.0	**0.1737**	0.010	5.7	ug/L	2	Standard
	Pb	206	284182.0	0.3	**49.3998**	0.214	0.4	ug/L	240	Standard
	Pb	207	234533.7	1.0	**45.8828**	0.448	1.0	ug/L	198	Standard
	Pb	208	1023299.4	0.1	**50.6751**	0.295	0.6	ug/L	883	Standard
	U	238	6312.3	2.1	**0.8483**	0.014	1.7	ug/L	3	Standard
>	**Bi**	209	563477.4	0.7				ug/L	523984	Standard

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:02:57
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**28.8214**	24.963	86.6	mg/L	0	Standard
Mg	24	151.7	18.2	**0.0816**	0.027	33.5	mg/L	37	Standard
K	39	1206.7	9.1	**1.8839**	0.198	10.5	mg/L	38	Standard
Ca	43	230.0	38.1	**8.3188**	4.002	48.1	mg/L	10	Standard
Fe	54	27991.0	1.6	**17.6350**	0.523	3.0	mg/L	179	Standard
Fe	57	7366.8	1.7	**17.6771**	0.511	2.9	mg/L	335	Standard
Sc-1	45	81510.4	1.4				mg/L	61029	Standard
Cl	35	28359.2	2.3				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	4510.7	3.7				ug/L	1547	Standard
P	31	20567.3	1.1				ug/L	15286	Standard
S	34	2398.5	2.1				ug/L	1572	Standard
Sr	88	138.3	12.7				ug/L	73	Standard
C	12	220.0	4.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	113.3	35.7				mg/L	3	Standard
Dy	164	44981.5	2.0				mg/L	6	Standard
Ho-1	165	27182.7	2.2				mg/L	10	Standard
Er	166	24606.7	4.9				mg/L	13	Standard
I	127	24668.4	3.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		128.648	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		116.023	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:02:57
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	113.109
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	107.537
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:02:57
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138101
Sample Date/Time: Wednesday, June 29, 2016 19:03:51
Number of Replicates: 3
Autosampler Position: 253
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89738.3	1.4				ug/L	74699	Standard
	Be	9	41.7	6.9	**0.0470**	0.003	6.5	ug/L	8	Standard
	Al	27	505536.2	3.8	**2.8498**	0.087	3.1	ug/L	1493	Standard
	Sc	45	72476.3	1.0				ug/L	61029	Standard
	Ti	47	238.7	3.3	**0.4178**	0.015	3.5	ug/L	35	Standard
	V	51	6994.4	2.1	**0.5793**	0.021	3.7	ug/L	1435	Standard
	Cr	52	13067.6	0.8	**0.4515**	0.011	2.5	ug/L	7635	Standard
	Cr	53	1060.0	5.3	**0.6298**	0.052	8.3	ug/L	363	Standard
	Mn	55	117506.1	2.5	**13.9721**	0.301	2.2	ug/L	747	Standard
	Co	59	2633.2	1.9	**0.2964**	0.006	2.1	ug/L	234	Standard
	Ni	60	1579.1	3.6	**0.5819**	0.030	5.1	ug/L	215	Standard
	Cu	65	1415.4	1.9	**0.5327**	0.017	3.2	ug/L	104	Standard
	Zn	66	4415.0	0.5	**2.9272**	0.025	0.9	ug/L	363	Standard
>	**Ge**	72	591804.3	0.7				ug/L	495142	Standard
	As	75	188.6	12.3	**0.2004**	0.017	8.4	ug/L	-50	Standard
	Se	82	17.7	38.3	**-0.0030**	0.051	1669.1	ug/L	14	Standard
	Se-1	77	53.3	24.5	**-0.0120**	0.145	1206.4	ug/L	42	Standard
>	**Ga**	71	651.7	3.8				mg/L	42	Standard
	Rb	85	6646.5	2.9				ug/L	17	Standard
	Y	89	496582.4	1.4				ug/L	417098	Standard
>	**Rh**	103	10.0	86.6				ug/L	13	Standard
	Mo	98	63.7	13.0	**0.0149**	0.003	17.4	ug/L	11	Standard
	Ag	107	138.0	7.0	**0.0054**	0.001	21.7	ug/L	68	Standard
	Cd	111	41.6	25.2	**0.0093**	0.004	47.8	mg/L	2	Standard
	Cd	114	78.4	6.5	**0.0111**	0.001	8.0	ug/L	12	Standard
>	**In**	115	680653.4	0.9				ug/L	587441	Standard
	Sn	118	558.3	7.0	**-0.0122**	0.005	43.2	ug/L	497	Standard
	Sb	123	143.1	16.8	**0.0222**	0.004	19.8	ug/L	78	Standard
	Ba	135	14129.9	0.6	**5.9423**	0.035	0.6	ug/L	24	Standard
	Ce	140	36747.8	1.4				ug/L	25	Standard
>	**Tb**	159	1046560.0	0.5				ug/L	939701	Standard
	Ho	165	620.0	6.6				ug/L	10	Standard
	Tl	203	40.7	9.9	**0.0027**	0.000	16.9	ug/L	4	Standard
	Tl	205	38.3	19.9	**0.0050**	0.001	18.8	ug/L	2	Standard
	Pb	206	6390.7	0.3	**1.0069**	0.002	0.2	ug/L	240	Standard
	Pb	207	5305.6	1.7	**0.9374**	0.017	1.8	ug/L	198	Standard
	Pb	208	22358.8	1.4	**1.0005**	0.020	2.0	ug/L	883	Standard
	U	238	147.3	10.0	**0.0149**	0.002	12.5	ug/L	3	Standard
>	**Bi**	209	581458.5	0.4				ug/L	523984	Standard

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:06:08
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	36.7	41.7	**-0.0251**	0.017	67.5	mg/L	37	Standard
K	39	73.3	27.6	**0.0225**	0.038	170.4	mg/L	38	Standard
Ca	43	25.0	40.0	**-1.1927**	0.538	45.1	mg/L	10	Standard
Fe	54	777.1	4.9	**0.4166**	0.030	7.2	mg/L	179	Standard
Fe	57	423.3	1.8	**0.1373**	0.012	9.1	mg/L	335	Standard
> **Sc-1**	45	72476.3	1.0				mg/L	61029	Standard
Cl	35	26907.2	2.4				ug/L	20195	Standard
Kr	83	1.7	124.9				ug/L	4	Standard
Br	81	2140.2	2.1				ug/L	1547	Standard
P	31	19854.7	4.5				ug/L	15286	Standard
S	34	2031.8	8.8				ug/L	1572	Standard
Sr	88	61.7	24.8				ug/L	73	Standard
C	12	66.7	82.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	1031.0	18.9				mg/L	6	Standard
Ho-1	165	620.0	6.6				mg/L	10	Standard
Er	166	536.7	12.4				mg/L	13	Standard
I	127	3175.3	27.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		120.134	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		119.522	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138101
Report Date/Time: Wednesday, June 29, 2016 19:06:08
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		115.868
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.969
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138102
Sample Date/Time: Wednesday, June 29, 2016 19:07:03
Number of Replicates: 3
Autosampler Position: 254
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	93009.7	1.6				ug/L	74699	Standard
	Be	9	1248.4	3.0	**1.1470**	0.051	4.5	ug/L	8	Standard
	Al	27	40124336.3	0.4	**219.2936**	2.793	1.3	ug/L	1493	Standard
	Sc	45	83437.5	1.7				ug/L	61029	Standard
	Ti	47	9733.8	0.5	**22.4436**	0.290	1.3	ug/L	35	Standard
	V	51	219254.4	2.3	**25.2954**	0.490	1.9	ug/L	1435	Standard
	Cr	52	167385.2	0.7	**18.8093**	0.105	0.6	ug/L	7635	Standard
	Cr	53	20240.2	3.5	**18.8976**	0.529	2.8	ug/L	363	Standard
	Mn	55	6543259.4	1.4	**825.1579**	15.855	1.9	ug/L	747	Standard
	Co	59	86492.3	2.3	**11.4640**	0.252	2.2	ug/L	234	Standard
	Ni	60	58233.4	1.3	**27.1864**	0.121	0.4	ug/L	215	Standard
	Cu	65	35144.1	1.8	**16.4764**	0.210	1.3	ug/L	104	Standard
	Zn	66	61921.6	1.7	**49.3206**	0.509	1.0	ug/L	363	Standard
>	**Ge**	72	563884.1	0.9				ug/L	495142	Standard
	As	75	5876.5	0.9	**4.6669**	0.080	1.7	ug/L	-50	Standard
	Se	82	62.7	20.2	**0.3638**	0.101	27.7	ug/L	14	Standard
	Se-1	77	116.3	8.0	**0.7757**	0.099	12.7	ug/L	42	Standard
>	**Ga**	71	24224.4	1.6				mg/L	42	Standard
	Rb	85	231886.4	1.9				ug/L	17	Standard
	Y	89	788603.8	0.2				ug/L	417098	Standard
>	**Rh**	103	76.7	50.7				ug/L	13	Standard
	Mo	98	590.2	6.7	**0.1909**	0.011	6.0	ug/L	11	Standard
	Ag	107	590.3	14.1	**0.0675**	0.010	15.2	ug/L	68	Standard
	Cd	111	402.0	10.7	**0.1627**	0.016	9.9	mg/L	2	Standard
	Cd	114	896.6	5.0	**0.1531**	0.007	4.5	ug/L	12	Standard
>	**In**	115	661468.7	1.5				ug/L	587441	Standard
	Sn	118	563.3	17.0	**-0.0090**	0.016	173.2	ug/L	497	Standard
	Sb	123	168.4	11.6	**0.0279**	0.004	15.0	ug/L	78	Standard
	Ba	135	766980.0	0.9	**332.7995**	3.011	0.9	ug/L	24	Standard
	Ce	140	1513996.2	0.9				ug/L	25	Standard
>	**Tb**	159	1057537.7	1.8				ug/L	939701	Standard
	Ho	165	25513.1	1.6				ug/L	10	Standard
	Tl	203	1321.7	3.8	**0.1489**	0.005	3.2	ug/L	4	Standard
	Tl	205	1201.7	6.5	**0.1579**	0.009	5.7	ug/L	2	Standard
	Pb	206	69846.7	2.1	**12.2126**	0.140	1.1	ug/L	240	Standard
	Pb	207	56032.1	2.8	**11.0205**	0.188	1.7	ug/L	198	Standard
	Pb	208	243026.5	0.6	**12.1032**	0.059	0.5	ug/L	883	Standard
	U	238	7981.8	0.7	**1.0851**	0.004	0.3	ug/L	3	Standard
>	**Bi**	209	557709.9	1.1				ug/L	523984	Standard

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:09:20
Page 1

Approved: July 04, 2016

Na	23	6.7	86.6	**56.4197**	48.874	86.6	mg/L	0	Standard
Mg	24	238.3	32.7	**0.1607**	0.077	48.1	mg/L	37	Standard
K	39	688.3	8.8	**0.9980**	0.083	8.3	mg/L	38	Standard
Ca	43	911.7	12.3	**39.6249**	5.390	13.6	mg/L	10	Standard
Fe	54	20800.6	0.9	**12.7625**	0.173	1.4	mg/L	179	Standard
Fe	57	6012.9	2.1	**13.8749**	0.248	1.8	mg/L	335	Standard
Sc-1	45	83437.5	1.7				mg/L	61029	Standard
Cl	35	29332.4	1.9				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2603.6	11.5				ug/L	1547	Standard
P	31	21632.1	2.1				ug/L	15286	Standard
S	34	2746.9	3.7				ug/L	1572	Standard
Sr	88	253.3	10.9				ug/L	73	Standard
C	12	116.7	27.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	33.3	45.8				mg/L	3	Standard
Dy	164	42471.8	1.0				mg/L	6	Standard
Ho-1	165	25513.1	1.6				mg/L	10	Standard
Er	166	22212.9	0.4				mg/L	13	Standard
I	127	23259.5	0.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		124.513	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		113.883	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:09:20
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.602
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.436
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:09:20
Page 3

Approved: July 04, 2016

Ba 135 Upper, S, EEE	Ba	135

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138102
Sample Date/Time: Wednesday, June 29, 2016 19:10:14
Number of Replicates: 3
Autosampler Position: 255
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85837.5	1.0				ug/L	74699	Standard
	Be	9	38.3	19.9	**0.0455**	0.007	16.1	ug/L	8	Standard
	Al	27	829500.0	3.1	**4.8979**	0.105	2.1	ug/L	1493	Standard
	Sc	45	72806.2	1.9				ug/L	61029	Standard
	Ti	47	228.0	6.2	**0.3999**	0.033	8.1	ug/L	35	Standard
	V	51	5937.2	2.5	**0.4694**	0.017	3.6	ug/L	1435	Standard
	Cr	52	11356.6	1.4	**0.2725**	0.020	7.2	ug/L	7635	Standard
	Cr	53	790.0	13.9	**0.3928**	0.096	24.5	ug/L	363	Standard
	Mn	55	140399.5	1.1	**16.9108**	0.252	1.5	ug/L	747	Standard
	Co	59	1986.1	4.2	**0.2171**	0.011	4.8	ug/L	234	Standard
	Ni	60	1398.7	1.4	**0.5081**	0.009	1.8	ug/L	215	Standard
	Cu	65	887.4	6.1	**0.2997**	0.027	8.9	ug/L	104	Standard
	Zn	66	3530.4	0.8	**2.2796**	0.035	1.5	ug/L	363	Standard
>	**Ge**	72	585343.3	0.6				ug/L	495142	Standard
	As	75	60.2	40.7	**0.1052**	0.019	17.9	ug/L	-50	Standard
	Se	82	19.5	23.7	**0.0127**	0.037	290.2	ug/L	14	Standard
	Se-1	77	54.0	19.3	**0.0032**	0.122	3775.1	ug/L	42	Standard
>	**Ga**	71	611.7	10.2				mg/L	42	Standard
	Rb	85	4939.1	4.6				ug/L	17	Standard
	Y	89	509652.0	1.0				ug/L	417098	Standard
>	**Rh**	103	5.0	100.0				ug/L	13	Standard
	Mo	98	44.1	7.7	**0.0084**	0.001	14.8	ug/L	11	Standard
	Ag	107	136.0	8.3	**0.0049**	0.002	33.6	ug/L	68	Standard
	Cd	111	13.6	15.2	**-0.0024**	0.001	32.7	mg/L	2	Standard
	Cd	114	34.6	39.9	**0.0036**	0.002	62.3	ug/L	12	Standard
>	**In**	115	689569.5	1.2				ug/L	587441	Standard
	Sn	118	635.0	8.5	**-0.0021**	0.008	369.1	ug/L	497	Standard
	Sb	123	120.9	7.4	**0.0179**	0.002	10.7	ug/L	78	Standard
	Ba	135	16312.1	1.7	**6.7731**	0.043	0.6	ug/L	24	Standard
	Ce	140	31461.0	1.2				ug/L	25	Standard
>	**Tb**	159	1048084.0	0.7				ug/L	939701	Standard
	Ho	165	503.3	20.9				ug/L	10	Standard
	Tl	203	28.0	9.4	**0.0013**	0.000	21.0	ug/L	4	Standard
	Tl	205	13.3	21.7	**0.0019**	0.000	19.6	ug/L	2	Standard
	Pb	206	1789.4	1.9	**0.2333**	0.006	2.6	ug/L	240	Standard
	Pb	207	1416.1	2.7	**0.2012**	0.009	4.3	ug/L	198	Standard
	Pb	208	6428.2	1.8	**0.2375**	0.004	1.7	ug/L	883	Standard
	U	238	157.3	5.4	**0.0164**	0.001	7.5	ug/L	3	Standard
>	**Bi**	209	576421.5	0.5				ug/L	523984	Standard

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:12:31
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	65.0	13.3	**0.0052**	0.008	161.5	mg/L	37	Standard
K	39	56.7	10.2	**-0.0092**	0.009	98.8	mg/L	38	Standard
Ca	43	46.7	16.4	**-0.0469**	0.452	963.7	mg/L	10	Standard
Fe	54	615.5	5.2	**0.2993**	0.025	8.2	mg/L	179	Standard
Fe	57	410.0	10.0	**0.0928**	0.096	103.4	mg/L	335	Standard
> **Sc-1**	45	72806.2	1.9				mg/L	61029	Standard
Cl	35	29355.1	3.4				ug/L	20195	Standard
Kr	83	2.3	24.7				ug/L	4	Standard
Br	81	1910.1	15.3				ug/L	1547	Standard
P	31	20493.9	2.3				ug/L	15286	Standard
S	34	2463.5	6.9				ug/L	1572	Standard
Sr	88	83.3	42.1				ug/L	73	Standard
C	12	50.0	34.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	991.0	1.5				mg/L	6	Standard
Ho-1	165	503.3	20.9				mg/L	10	Standard
Er	166	466.7	4.5				mg/L	13	Standard
I	127	2401.9	23.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		114.911	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		118.217	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138102
Report Date/Time: Wednesday, June 29, 2016 19:12:31
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.385
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.007
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 19:13:27
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	83999.1	3.6				ug/L	74699	Standard
	Be	9	57157.7	3.6	**57.7443**	0.451	0.8	ug/L	8	Standard
	Al	27	8858742.4	2.0	**53.6133**	0.921	1.7	ug/L	1493	Standard
	Sc	45	70122.0	3.5				ug/L	61029	Standard
	Ti	47	47445.5	0.9	**109.3528**	3.353	3.1	ug/L	35	Standard
	V	51	471838.1	3.0	**54.4025**	0.466	0.9	ug/L	1435	Standard
	Cr	52	466548.5	1.7	**54.0130**	1.261	2.3	ug/L	7635	Standard
	Cr	53	57204.5	1.8	**53.7881**	1.637	3.0	ug/L	363	Standard
	Mn	55	422873.4	1.8	**52.9497**	1.131	2.1	ug/L	747	Standard
	Co	59	420575.5	1.8	**55.6399**	1.134	2.0	ug/L	234	Standard
	Ni	60	114954.2	1.8	**53.5507**	1.121	2.1	ug/L	215	Standard
	Cu	65	113826.7	2.0	**53.3592**	1.065	2.0	ug/L	104	Standard
	Zn	66	66150.5	1.8	**52.4882**	1.217	2.3	ug/L	363	Standard
>	Ge	72	566689.7	3.9				ug/L	495142	Standard
	As	75	66474.1	0.9	**51.9549**	1.668	3.2	ug/L	-50	Standard
	Se	82	6669.2	0.5	**53.0020**	1.808	3.4	ug/L	14	Standard
	Se-1	77	4569.4	1.3	**54.0883**	1.963	3.6	ug/L	42	Standard
>	Ga	71	60.0	25.0				mg/L	42	Standard
	Rb	85	973.4	4.9				ug/L	17	Standard
	Y	89	482055.4	0.8				ug/L	417098	Standard
>	Rh	103	21.7	35.3				ug/L	13	Standard
	Mo	98	311127.6	1.9	**101.8685**	1.475	1.4	ug/L	11	Standard
	Ag	107	381455.1	1.8	**51.0679**	0.625	1.2	ug/L	68	Standard
	Cd	111	122240.5	1.1	**51.0178**	0.254	0.5	mg/L	2	Standard
	Cd	114	297421.0	1.0	**50.6157**	0.803	1.6	ug/L	12	Standard
>	In	115	673075.7	0.6				ug/L	587441	Standard
	Sn	118	341770.5	1.5	**50.9666**	0.480	0.9	ug/L	497	Standard
	Sb	123	266115.5	1.9	**49.7407**	0.724	1.5	ug/L	78	Standard
	Ba	135	114468.8	0.7	**48.7952**	0.078	0.2	ug/L	24	Standard
	Ce	140	298.3	40.9				ug/L	25	Standard
>	Tb	159	1044481.4	1.5				ug/L	939701	Standard
	Ho	165	18.3	83.3				ug/L	10	Standard
	Tl	203	467319.7	1.5	**52.0663**	0.128	0.2	ug/L	4	Standard
	Tl	205	421316.6	1.5	**54.0865**	0.384	0.7	ug/L	2	Standard
	Pb	206	294059.2	2.1	**50.4836**	0.259	0.5	ug/L	240	Standard
	Pb	207	260305.8	2.4	**50.2976**	0.406	0.8	ug/L	198	Standard
	Pb	208	1086866.5	0.8	**53.1659**	0.550	1.0	ug/L	883	Standard
	U	238	395799.2	0.7	**52.8135**	0.760	1.4	ug/L	3	Standard
>	Bi	209	570529.8	1.7				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:15:44
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4777.4	3.9	**5.3091**	0.347	6.5	mg/L	37	Standard
K	39	2250.2	3.7	**4.2173**	0.198	4.7	mg/L	38	Standard
Ca	43	128.3	12.5	**4.5532**	1.080	23.7	mg/L	10	Standard
Fe	54	7377.7	1.8	**5.3086**	0.144	2.7	mg/L	179	Standard
Fe	57	2055.1	1.5	**5.0082**	0.127	2.5	mg/L	335	Standard
> **Sc-1**	45	70122.0	3.5				mg/L	61029	Standard
Cl	35	30001.0	0.9				ug/L	20195	Standard
Kr	83	1.7	91.7				ug/L	4	Standard
Br	81	2183.5	6.8				ug/L	1547	Standard
P	31	22995.8	1.5				ug/L	15286	Standard
S	34	2830.3	2.9				ug/L	1572	Standard
Sr	88	66.7	24.1				ug/L	73	Standard
C	12	80.0	50.0				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	9.0	108.0				mg/L	6	Standard
Ho-1	165	18.3	83.3				mg/L	10	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	1968.5	24.7				mg/L	1392	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	115.489		
	Al	27	107.227		
	Sc	45			
	Ti	47	109.353		
	V	51	108.805		
	Cr	52	108.026		
	Cr	53			
	Mn	55	105.899		
	Co	59	111.280		
	Ni	60	107.101		
	Cu	65	106.718		
	Zn	66	104.976		
>	Ge	72		114.450	
	As	75	103.910		
	Se	82	106.004		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:15:44
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.868		
Ag	107	102.136		
Cd	111	102.036		
Cd	114			
In	115		114.578	
Sn	118	101.933		
Sb	123	99.481		
Ba	135	97.590		
Ce	140			
Tb	159			
Ho	165			
Tl	203	104.133		
Tl	205			
Pb	206			
Pb	207			
Pb	208	106.332		
U	238	105.627		
Bi	209		108.883	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:15:44
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 19:16:39
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	83692.3	2.8				ug/L	74699	Standard
	Be	9	18.3	15.7	**0.0263**	0.003	12.9	ug/L	8	Standard
	Al	27	3213.7	25.4	**0.0061**	0.004	71.9	ug/L	1493	Standard
	Sc	45	70726.5	3.0				ug/L	61029	Standard
	Ti	47	41.0	12.2	**-0.0158**	0.012	73.1	ug/L	35	Standard
	V	51	1525.1	1.9	**-0.0214**	0.005	22.4	ug/L	1435	Standard
	Cr	52	7659.9	0.9	**-0.1324**	0.015	11.0	ug/L	7635	Standard
	Cr	53	410.0	16.0	**0.0526**	0.061	116.5	ug/L	363	Standard
	Mn	55	997.7	7.3	**-0.0244**	0.010	41.0	ug/L	747	Standard
	Co	59	260.3	1.6	**-0.0033**	0.000	10.0	ug/L	234	Standard
	Ni	60	195.0	3.6	**-0.0341**	0.004	11.2	ug/L	215	Standard
	Cu	65	140.7	10.7	**-0.0385**	0.007	19.0	ug/L	104	Standard
	Zn	66	386.7	3.7	**-0.1493**	0.008	5.7	ug/L	363	Standard
>	**Ge**	72	574185.4	0.9				ug/L	495142	Standard
	As	75	-49.3	11.6	**0.0216**	0.004	19.9	ug/L	-50	Standard
	Se	82	18.5	26.4	**0.0072**	0.039	544.3	ug/L	14	Standard
	Se-1	77	51.3	18.1	**-0.0155**	0.115	739.5	ug/L	42	Standard
>	**Ga**	71	41.7	30.2				mg/L	42	Standard
	Rb	85	33.3	22.9				ug/L	17	Standard
	Y	89	508899.9	2.4				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	287.2	24.5	**0.0857**	0.020	22.8	ug/L	11	Standard
	Ag	107	140.0	24.8	**0.0054**	0.004	76.0	ug/L	68	Standard
	Cd	111	18.8	30.5	**-0.0003**	0.002	828.9	mg/L	2	Standard
	Cd	114	42.6	5.3	**0.0050**	0.000	2.5	ug/L	12	Standard
>	**In**	115	688571.2	3.6				ug/L	587441	Standard
	Sn	118	896.7	13.9	**0.0365**	0.020	55.6	ug/L	497	Standard
	Sb	123	419.7	28.4	**0.0731**	0.025	33.8	ug/L	78	Standard
	Ba	135	58.0	33.3	**0.0081**	0.008	93.7	ug/L	24	Standard
	Ce	140	43.3	17.6				ug/L	25	Standard
>	**Tb**	159	1032903.3	2.1				ug/L	939701	Standard
	Ho	165	11.7	24.7				ug/L	10	Standard
	Tl	203	69.3	37.1	**0.0059**	0.003	47.0	ug/L	4	Standard
	Tl	205	78.3	57.2	**0.0101**	0.006	55.2	ug/L	2	Standard
	Pb	206	303.3	6.2	**-0.0195**	0.003	14.1	ug/L	240	Standard
	Pb	207	260.7	8.9	**-0.0200**	0.005	24.1	ug/L	198	Standard
	Pb	208	1157.4	6.9	**-0.0179**	0.003	17.6	ug/L	883	Standard
	U	238	52.0	46.8	**0.0024**	0.003	129.1	ug/L	3	Standard
>	**Bi**	209	575232.2	1.3				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:18:56
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	36.7	28.4	**-0.0241**	0.012	50.4	mg/L	37	Standard
K	39	50.0	26.5	**-0.0190**	0.023	121.8	mg/L	38	Standard
Ca	43	26.7	39.0	**-1.0625**	0.584	55.0	mg/L	10	Standard
Fe	54	147.5	7.0	**-0.0305**	0.005	15.7	mg/L	179	Standard
Fe	57	216.7	8.1	**-0.4394**	0.033	7.6	mg/L	335	Standard
Sc-1	45	70726.5	3.0				mg/L	61029	Standard
Cl	35	29273.6	1.2				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2073.5	10.2				ug/L	1547	Standard
P	31	23289.7	10.5				ug/L	15286	Standard
S	34	2535.2	3.8				ug/L	1572	Standard
Sr	88	68.3	37.5				ug/L	73	Standard
C	12	66.7	22.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	16.3	126.2				mg/L	6	Standard
Ho-1	165	11.7	24.7				mg/L	10	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	1766.8	31.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.964	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:18:56
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.215
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.780
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:18:56
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138103
Sample Date/Time: Wednesday, June 29, 2016 19:19:52
Number of Replicates: 3
Autosampler Position: 256
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	88921.9	2.3				ug/L	74699	Standard
	Be	9	1191.7	0.9	**1.1451**	0.025	2.2	ug/L	8	Standard
	Al	27	19716870.1	2.2	**112.7701**	5.073	4.5	ug/L	1493	Standard
	Sc	45	82006.5	3.2				ug/L	61029	Standard
	Ti	47	8807.2	1.6	**20.0178**	0.269	1.3	ug/L	35	Standard
	V	51	208539.0	1.3	**23.7197**	0.153	0.6	ug/L	1435	Standard
	Cr	52	166877.0	1.1	**18.4795**	0.081	0.4	ug/L	7635	Standard
	Cr	53	19956.5	0.7	**18.3731**	0.126	0.7	ug/L	363	Standard
	Mn	55	4948296.7	1.6	**615.4227**	6.037	1.0	ug/L	747	Standard
	Co	59	82754.8	2.1	**10.8165**	0.154	1.4	ug/L	234	Standard
	Ni	60	51698.7	0.5	**23.7942**	0.298	1.3	ug/L	215	Standard
	Cu	65	48855.0	1.6	**22.6321**	0.340	1.5	ug/L	104	Standard
	Zn	66	105467.2	1.2	**83.1726**	0.384	0.5	ug/L	363	Standard
>	**Ge**	72	571655.8	0.7				ug/L	495142	Standard
	As	75	9636.9	0.9	**7.5116**	0.042	0.6	ug/L	-50	Standard
	Se	82	70.5	8.3	**0.4184**	0.042	10.1	ug/L	14	Standard
	Se-1	77	129.0	9.9	**0.9077**	0.153	16.9	ug/L	42	Standard
>	**Ga**	71	22341.5	2.4				mg/L	42	Standard
	Rb	85	210691.8	1.9				ug/L	17	Standard
	Y	89	776017.6	0.9				ug/L	417098	Standard
>	**Rh**	103	61.7	24.8				ug/L	13	Standard
	Mo	98	1324.4	9.6	**0.4208**	0.045	10.6	ug/L	11	Standard
	Ag	107	918.4	2.4	**0.1080**	0.004	3.7	ug/L	68	Standard
	Cd	111	891.1	4.2	**0.3578**	0.020	5.5	mg/L	2	Standard
	Cd	114	2180.6	6.0	**0.3625**	0.017	4.8	ug/L	12	Standard
>	**In**	115	684819.6	1.2				ug/L	587441	Standard
	Sn	118	1785.1	3.5	**0.1674**	0.006	3.8	ug/L	497	Standard
	Sb	123	354.4	3.1	**0.0609**	0.003	4.6	ug/L	78	Standard
	Ba	135	461342.5	1.4	**193.3434**	2.631	1.4	ug/L	24	Standard
	Ce	140	1368480.3	1.7				ug/L	25	Standard
>	**Tb**	159	1075718.6	2.2				ug/L	939701	Standard
	Ho	165	22026.0	1.7				ug/L	10	Standard
	Tl	203	1135.0	2.0	**0.1247**	0.001	0.6	ug/L	4	Standard
	Tl	205	1093.4	11.6	**0.1404**	0.014	10.2	ug/L	2	Standard
	Pb	206	147882.6	1.3	**25.3549**	0.415	1.6	ug/L	240	Standard
	Pb	207	119862.3	0.6	**23.1268**	0.424	1.8	ug/L	198	Standard
	Pb	208	520445.5	2.0	**25.4242**	0.898	3.5	ug/L	883	Standard
	U	238	4663.1	0.9	**0.6178**	0.006	0.9	ug/L	3	Standard
>	**Bi**	209	570570.9	1.5				ug/L	523984	Standard

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:22:09
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	126.7	27.7	**0.0563**	0.032	56.3	mg/L	37	Standard
K	39	878.4	3.1	**1.3326**	0.091	6.9	mg/L	38	Standard
Ca	43	426.7	4.7	**17.5293**	0.460	2.6	mg/L	10	Standard
Fe	54	21094.2	3.4	**13.1896**	0.865	6.6	mg/L	179	Standard
Fe	57	5826.1	1.4	**13.6721**	0.491	3.6	mg/L	335	Standard
> **Sc-1**	45	82006.5	3.2				mg/L	61029	Standard
Cl	35	28772.6	2.0				ug/L	20195	Standard
Kr	83	3.0	66.7				ug/L	4	Standard
Br	81	3580.4	3.6				ug/L	1547	Standard
P	31	21750.6	1.3				ug/L	15286	Standard
S	34	2703.6	0.6				ug/L	1572	Standard
Sr	88	178.3	13.8				ug/L	73	Standard
C	12	233.3	9.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	96.7	36.3				mg/L	3	Standard
Dy	164	37298.5	1.0				mg/L	6	Standard
Ho-1	165	22026.0	1.7				mg/L	10	Standard
Er	166	20383.7	1.3				mg/L	13	Standard
I	127	19701.1	3.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		119.041	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.453	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:22:09
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.577
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.891
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:22:09
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138103
Sample Date/Time: Wednesday, June 29, 2016 19:23:03
Number of Replicates: 3
Autosampler Position: 257
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82351.7	3.3				ug/L	74699	Standard
	Be	9	43.3	24.0	**0.0523**	0.011	20.3	ug/L	8	Standard
	Al	27	430174.7	3.0	**2.6419**	0.016	0.6	ug/L	1493	Standard
	Sc	45	73040.6	1.2				ug/L	61029	Standard
	Ti	47	215.7	6.2	**0.3839**	0.033	8.6	ug/L	35	Standard
	V	51	5937.3	2.7	**0.4853**	0.017	3.5	ug/L	1435	Standard
	Cr	52	11052.0	1.3	**0.2679**	0.017	6.2	ug/L	7635	Standard
	Cr	53	723.4	5.6	**0.3480**	0.038	11.0	ug/L	363	Standard
	Mn	55	109527.8	0.6	**13.4774**	0.102	0.8	ug/L	747	Standard
	Co	59	1998.1	1.8	**0.2248**	0.006	2.5	ug/L	234	Standard
	Ni	60	1283.7	2.7	**0.4700**	0.016	3.4	ug/L	215	Standard
	Cu	65	1177.7	1.9	**0.4444**	0.009	2.1	ug/L	104	Standard
	Zn	66	4258.3	1.3	**2.9218**	0.029	1.0	ug/L	363	Standard
>	**Ge**	72	571680.5	0.6				ug/L	495142	Standard
	As	75	130.4	5.6	**0.1604**	0.006	3.7	ug/L	-50	Standard
	Se	82	14.1	16.1	**-0.0268**	0.018	65.5	ug/L	14	Standard
	Se-1	77	58.0	6.9	**0.0653**	0.044	66.8	ug/L	42	Standard
>	**Ga**	71	521.7	4.0				mg/L	42	Standard
	Rb	85	4554.0	2.0				ug/L	17	Standard
	Y	89	512515.9	2.6				ug/L	417098	Standard
>	**Rh**	103	11.7	24.7				ug/L	13	Standard
	Mo	98	90.4	4.4	**0.0229**	0.001	6.5	ug/L	11	Standard
	Ag	107	144.3	4.6	**0.0058**	0.001	16.1	ug/L	68	Standard
	Cd	111	21.8	34.6	**0.0009**	0.003	335.6	mg/L	2	Standard
	Cd	114	67.8	46.0	**0.0090**	0.005	55.6	ug/L	12	Standard
>	**In**	115	696556.9	1.5				ug/L	587441	Standard
	Sn	118	640.0	3.6	**-0.0023**	0.005	204.9	ug/L	497	Standard
	Sb	123	182.0	15.1	**0.0287**	0.006	19.2	ug/L	78	Standard
	Ba	135	9992.3	2.7	**4.1009**	0.069	1.7	ug/L	24	Standard
	Ce	140	29587.2	2.0				ug/L	25	Standard
>	**Tb**	159	1058949.6	2.5				ug/L	939701	Standard
	Ho	165	493.3	6.5				ug/L	10	Standard
	Tl	203	41.7	21.8	**0.0028**	0.001	37.1	ug/L	4	Standard
	Tl	205	43.3	54.5	**0.0057**	0.003	52.4	ug/L	2	Standard
	Pb	206	3583.8	1.2	**0.5358**	0.012	2.2	ug/L	240	Standard
	Pb	207	2855.9	1.4	**0.4744**	0.011	2.4	ug/L	198	Standard
	Pb	208	12510.0	0.4	**0.5295**	0.006	1.1	ug/L	883	Standard
	U	238	110.3	6.2	**0.0101**	0.001	8.0	ug/L	3	Standard
>	**Bi**	209	579224.4	1.1				ug/L	523984	Standard

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:25:20
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	51.7	29.6	**-0.0092**	0.017	184.4	mg/L	37	Standard
K	39	41.7	61.6	**-0.0375**	0.047	124.3	mg/L	38	Standard
Ca	43	26.7	10.8	**-1.1162**	0.146	13.1	mg/L	10	Standard
Fe	54	584.0	15.6	**0.2752**	0.062	22.7	mg/L	179	Standard
Fe	57	388.3	5.2	**0.0289**	0.063	219.5	mg/L	335	Standard
> **Sc-1**	45	73040.6	1.2				mg/L	61029	Standard
Cl	35	28933.6	1.8				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2263.5	5.9				ug/L	1547	Standard
P	31	20116.7	3.4				ug/L	15286	Standard
S	34	2530.2	7.0				ug/L	1572	Standard
Sr	88	75.0	23.1				ug/L	73	Standard
C	12	66.7	56.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	963.7	4.1				mg/L	6	Standard
Ho-1	165	493.3	6.5				mg/L	10	Standard
Er	166	410.0	6.5				mg/L	13	Standard
I	127	2373.5	31.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.245	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.458	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138103
Report Date/Time: Wednesday, June 29, 2016 19:25:20
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		118.575
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.542
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138104
Sample Date/Time: Wednesday, June 29, 2016 19:26:15
Number of Replicates: 3
Autosampler Position: 258
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	85261.0	2.3				ug/L	74699	Standard
	Be	9	1065.0	16.2	**1.0661**	0.158	14.8	ug/L	8	Standard
	Al	27	15142099.3	1.2	**90.2750**	1.228	1.4	ug/L	1493	Standard
	Sc	45	77933.3	2.2				ug/L	61029	Standard
	Ti	47	7118.0	1.5	**16.2707**	0.118	0.7	ug/L	35	Standard
	V	51	161044.2	0.8	**18.4034**	0.196	1.1	ug/L	1435	Standard
	Cr	52	146045.2	0.2	**16.1669**	0.325	2.0	ug/L	7635	Standard
	Cr	53	17793.8	1.2	**16.4664**	0.462	2.8	ug/L	363	Standard
	Mn	55	4299510.8	1.0	**538.5113**	8.104	1.5	ug/L	747	Standard
	Co	59	67815.8	1.0	**8.9211**	0.182	2.0	ug/L	234	Standard
	Ni	60	46536.4	1.1	**21.5551**	0.152	0.7	ug/L	215	Standard
	Cu	65	28919.6	1.2	**13.4506**	0.294	2.2	ug/L	104	Standard
	Zn	66	52781.1	1.6	**41.6917**	0.739	1.8	ug/L	363	Standard
>	**Ge**	72	567729.5	1.7				ug/L	495142	Standard
	As	75	4550.5	2.3	**3.6043**	0.141	3.9	ug/L	-50	Standard
	Se	82	43.5	41.5	**0.2091**	0.148	70.8	ug/L	14	Standard
	Se-1	77	118.3	4.0	**0.7906**	0.045	5.7	ug/L	42	Standard
>	**Ga**	71	21335.0	2.6				mg/L	42	Standard
	Rb	85	199689.0	1.1				ug/L	17	Standard
	Y	89	717025.3	1.7				ug/L	417098	Standard
>	**Rh**	103	53.3	39.0				ug/L	13	Standard
	Mo	98	750.2	3.3	**0.2455**	0.003	1.3	ug/L	11	Standard
	Ag	107	572.7	8.1	**0.0655**	0.005	7.5	ug/L	68	Standard
	Cd	111	313.3	5.6	**0.1259**	0.007	5.6	mg/L	2	Standard
	Cd	114	729.6	4.4	**0.1250**	0.007	5.3	ug/L	12	Standard
>	**In**	115	657920.1	2.1				ug/L	587441	Standard
	Sn	118	513.3	6.6	**-0.0162**	0.005	32.2	ug/L	497	Standard
	Sb	123	156.7	10.3	**0.0258**	0.004	14.3	ug/L	78	Standard
	Ba	135	278778.0	0.4	**121.6374**	2.867	2.4	ug/L	24	Standard
	Ce	140	1205012.6	0.9				ug/L	25	Standard
>	**Tb**	159	1028655.6	1.0				ug/L	939701	Standard
	Ho	165	19245.5	0.5				ug/L	10	Standard
	Tl	203	1022.7	4.2	**0.1168**	0.004	3.6	ug/L	4	Standard
	Tl	205	930.0	6.2	**0.1244**	0.008	6.1	ug/L	2	Standard
	Pb	206	75172.7	1.1	**13.3810**	0.239	1.8	ug/L	240	Standard
	Pb	207	60753.2	0.6	**12.1662**	0.152	1.3	ug/L	198	Standard
	Pb	208	268421.8	0.9	**13.6089**	0.053	0.4	ug/L	883	Standard
	U	238	4956.8	1.6	**0.6840**	0.014	2.1	ug/L	3	Standard
>	**Bi**	209	548179.9	0.7				ug/L	523984	Standard

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:28:31
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**29.7760**	51.565	173.2	mg/L	0	Standard
Mg	24	116.7	8.9	**0.0527**	0.008	15.2	mg/L	37	Standard
K	39	606.7	4.1	**0.9368**	0.065	7.0	mg/L	38	Standard
Ca	43	223.3	26.5	**8.5548**	3.074	35.9	mg/L	10	Standard
Fe	54	18758.0	3.4	**12.3136**	0.167	1.4	mg/L	179	Standard
Fe	57	5059.2	5.4	**12.3881**	0.534	4.3	mg/L	335	Standard
Sc-1	45	77933.3	2.2				mg/L	61029	Standard
Cl	35	27511.6	2.5				ug/L	20195	Standard
Kr	83	4.0	50.0				ug/L	4	Standard
Br	81	2450.2	6.4				ug/L	1547	Standard
P	31	19838.0	2.4				ug/L	15286	Standard
S	34	2438.5	4.0				ug/L	1572	Standard
Sr	88	130.0	10.2				ug/L	73	Standard
C	12	76.7	41.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	16.7	124.9				mg/L	3	Standard
Dy	164	31319.3	1.6				mg/L	6	Standard
Ho-1	165	19245.5	0.5				mg/L	10	Standard
Er	166	17009.5	2.0				mg/L	13	Standard
I	127	14255.0	0.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		114.140	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.660	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:28:31
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	111.998
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.618
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:28:31
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138104
Sample Date/Time: Wednesday, June 29, 2016 19:29:26
Number of Replicates: 3
Autosampler Position: 259
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81019.4	0.7				ug/L	74699	Standard
	Be	9	41.7	18.3	**0.0512**	0.008	15.3	ug/L	8	Standard
	Al	27	329954.8	1.9	**2.0568**	0.044	2.2	ug/L	1493	Standard
	Sc	45	70853.8	4.4				ug/L	61029	Standard
	Ti	47	170.0	4.8	**0.2650**	0.022	8.2	ug/L	35	Standard
	V	51	4919.9	1.5	**0.3475**	0.009	2.6	ug/L	1435	Standard
	Cr	52	11144.8	0.5	**0.2299**	0.010	4.3	ug/L	7635	Standard
	Cr	53	798.4	9.1	**0.3907**	0.073	18.6	ug/L	363	Standard
	Mn	55	93006.7	2.2	**10.9903**	0.351	3.2	ug/L	747	Standard
	Co	59	1669.1	4.9	**0.1734**	0.010	5.5	ug/L	234	Standard
	Ni	60	1239.4	3.4	**0.4280**	0.019	4.5	ug/L	215	Standard
	Cu	65	783.4	0.7	**0.2472**	0.006	2.3	ug/L	104	Standard
	Zn	66	3223.7	4.0	**2.0065**	0.120	6.0	ug/L	363	Standard
>	**Ge**	72	593953.7	1.0				ug/L	495142	Standard
	As	75	25.6	116.2	**0.0786**	0.022	28.0	ug/L	-50	Standard
	Se	82	15.5	11.3	**-0.0201**	0.014	71.9	ug/L	14	Standard
	Se-1	77	55.7	3.7	**0.0130**	0.019	145.3	ug/L	42	Standard
>	**Ga**	71	496.7	9.6				mg/L	42	Standard
	Rb	85	4318.9	1.6				ug/L	17	Standard
	Y	89	493099.4	1.2				ug/L	417098	Standard
>	**Rh**	103	8.3	91.7				ug/L	13	Standard
	Mo	98	63.8	10.3	**0.0154**	0.002	13.8	ug/L	11	Standard
	Ag	107	130.3	15.7	**0.0047**	0.003	56.8	ug/L	68	Standard
	Cd	111	7.2	21.0	**-0.0049**	0.001	12.8	mg/L	2	Standard
	Cd	114	27.0	44.7	**0.0025**	0.002	80.8	ug/L	12	Standard
>	**In**	115	665068.2	0.5				ug/L	587441	Standard
	Sn	118	545.0	15.7	**-0.0123**	0.013	107.4	ug/L	497	Standard
	Sb	123	150.6	9.1	**0.0243**	0.003	11.0	ug/L	78	Standard
	Ba	135	6015.5	0.5	**2.5800**	0.016	0.6	ug/L	24	Standard
	Ce	140	25678.4	3.6				ug/L	25	Standard
>	**Tb**	159	1023204.1	0.9				ug/L	939701	Standard
	Ho	165	420.0	5.5				ug/L	10	Standard
	Tl	203	37.3	22.8	**0.0024**	0.001	40.1	ug/L	4	Standard
	Tl	205	30.0	50.0	**0.0040**	0.002	46.9	ug/L	2	Standard
	Pb	206	1995.8	4.0	**0.2734**	0.016	6.0	ug/L	240	Standard
	Pb	207	1576.4	5.4	**0.2364**	0.019	8.0	ug/L	198	Standard
	Pb	208	6805.9	1.7	**0.2606**	0.003	1.1	ug/L	883	Standard
	U	238	99.3	2.5	**0.0089**	0.000	2.4	ug/L	3	Standard
>	**Bi**	209	568219.3	0.9				ug/L	523984	Standard

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:31:43
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	30.0	66.7	**-0.0312**	0.023	74.2	mg/L	37	Standard
K	39	65.0	30.8	**0.0108**	0.043	395.4	mg/L	38	Standard
Ca	43	20.0	0.0	**-1.4338**	0.047	3.3	mg/L	10	Standard
Fe	54	499.3	4.0	**0.2265**	0.010	4.2	mg/L	179	Standard
Fe	57	346.7	6.8	**-0.0592**	0.056	94.8	mg/L	335	Standard
> **Sc-1**	45	70853.8	4.4				mg/L	61029	Standard
Cl	35	26917.9	3.4				ug/L	20195	Standard
Kr	83	2.0	0.0				ug/L	4	Standard
Br	81	1996.8	2.8				ug/L	1547	Standard
P	31	19530.9	2.5				ug/L	15286	Standard
S	34	2235.2	2.4				ug/L	1572	Standard
Sr	88	70.0	18.9				ug/L	73	Standard
C	12	83.3	38.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	818.0	9.1				mg/L	6	Standard
Ho-1	165	420.0	5.5				mg/L	10	Standard
Er	166	390.0	11.8				mg/L	13	Standard
I	127	1930.1	32.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		108.461	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		119.956	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138104
Report Date/Time: Wednesday, June 29, 2016 19:31:43
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.214
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.442
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138105
Sample Date/Time: Wednesday, June 29, 2016 19:32:37
Number of Replicates: 3
Autosampler Position: 260
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	86630.3	1.6				ug/L	74699	Standard
	Be	9	1345.1	15.8	**1.3230**	0.188	14.2	ug/L	8	Standard
	Al	27	25518871.0	0.5	**149.7364**	2.025	1.4	ug/L	1493	Standard
	Sc	45	82793.9	1.1				ug/L	61029	Standard
	Ti	47	9047.0	1.1	**21.0343**	0.216	1.0	ug/L	35	Standard
	V	51	241144.8	2.1	**28.0828**	0.136	0.5	ug/L	1435	Standard
	Cr	52	185820.6	3.2	**21.1812**	0.417	2.0	ug/L	7635	Standard
	Cr	53	22667.0	3.7	**21.3886**	0.515	2.4	ug/L	363	Standard
	Mn	55	5666490.2	1.7	**720.7034**	2.244	0.3	ug/L	747	Standard
	Co	59	89984.1	1.6	**12.0322**	0.097	0.8	ug/L	234	Standard
	Ni	60	57928.9	2.2	**27.2782**	0.178	0.7	ug/L	215	Standard
	Cu	65	41127.1	3.0	**19.4650**	0.281	1.4	ug/L	104	Standard
	Zn	66	70831.6	3.1	**56.9699**	0.880	1.5	ug/L	363	Standard
>	**Ge**	72	559032.2	1.6				ug/L	495142	Standard
	As	75	9088.7	3.2	**7.2452**	0.123	1.7	ug/L	-50	Standard
	Se	82	66.3	23.6	**0.3960**	0.117	29.7	ug/L	14	Standard
	Se-1	77	130.0	4.3	**0.9538**	0.045	4.7	ug/L	42	Standard
>	**Ga**	71	25715.1	2.4				mg/L	42	Standard
	Rb	85	243499.9	2.1				ug/L	17	Standard
	Y	89	766854.2	0.5				ug/L	417098	Standard
>	**Rh**	103	51.7	5.6				ug/L	13	Standard
	Mo	98	760.8	5.6	**0.2496**	0.019	7.7	ug/L	11	Standard
	Ag	107	727.0	8.3	**0.0869**	0.010	11.3	ug/L	68	Standard
	Cd	111	420.9	5.2	**0.1721**	0.013	7.6	mg/L	2	Standard
	Cd	114	1021.1	15.6	**0.1761**	0.030	17.3	ug/L	12	Standard
>	**In**	115	657583.1	2.2				ug/L	587441	Standard
	Sn	118	706.7	9.4	**0.0135**	0.012	91.4	ug/L	497	Standard
	Sb	123	160.0	3.9	**0.0265**	0.002	7.0	ug/L	78	Standard
	Ba	135	489546.4	1.0	**213.7510**	6.703	3.1	ug/L	24	Standard
	Ce	140	1473172.3	0.9				ug/L	25	Standard
>	**Tb**	159	1043758.3	1.4				ug/L	939701	Standard
	Ho	165	23187.7	0.9				ug/L	10	Standard
	Tl	203	1287.4	7.3	**0.1458**	0.012	8.5	ug/L	4	Standard
	Tl	205	1168.4	5.5	**0.1545**	0.010	6.8	ug/L	2	Standard
	Pb	206	110607.4	2.3	**19.4839**	0.695	3.6	ug/L	240	Standard
	Pb	207	89635.8	2.2	**17.7667**	0.631	3.6	ug/L	198	Standard
	Pb	208	392577.0	0.9	**19.6965**	0.434	2.2	ug/L	883	Standard
	U	238	8119.8	0.8	**1.1095**	0.020	1.8	ug/L	3	Standard
>	**Bi**	209	554986.0	1.4				ug/L	523984	Standard

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:34:54
Page 1

Approved: July 04, 2016

Na	23	5.0	100.0	**42.5482**	42.299	99.4	mg/L	0	Standard
Mg	24	165.0	16.0	**0.0921**	0.027	29.3	mg/L	37	Standard
K	39	870.0	2.5	**1.3032**	0.021	1.6	mg/L	38	Standard
Ca	43	623.3	6.5	**26.5121**	2.122	8.0	mg/L	10	Standard
Fe	54	24836.5	3.4	**15.3836**	0.500	3.3	mg/L	179	Standard
Fe	57	6766.5	3.6	**15.8771**	0.513	3.2	mg/L	335	Standard
Sc-1	45	82793.9	1.1				mg/L	61029	Standard
Cl	35	27591.1	0.8				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	2510.2	7.0				ug/L	1547	Standard
P	31	20592.3	2.0				ug/L	15286	Standard
S	34	2640.2	1.7				ug/L	1572	Standard
Sr	88	220.0	6.0				ug/L	73	Standard
C	12	86.7	24.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	46.7	44.6				mg/L	3	Standard
Dy	164	40115.6	1.1				mg/L	6	Standard
Ho-1	165	23187.7	0.9				mg/L	10	Standard
Er	166	21645.5	1.0				mg/L	13	Standard
I	127	13629.5	5.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		115.973	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		112.903	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:34:54
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.940
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.916
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:34:54
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138105
Sample Date/Time: Wednesday, June 29, 2016 19:35:50
Number of Replicates: 3
Autosampler Position: 401
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81775.3	3.6				ug/L	74699	Standard
	Be	9	45.0	11.1	**0.0544**	0.006	11.5	ug/L	8	Standard
	Al	27	547945.2	2.8	**3.3933**	0.048	1.4	ug/L	1493	Standard
	Sc	45	73444.3	1.5				ug/L	61029	Standard
	Ti	47	212.7	8.9	**0.3795**	0.037	9.9	ug/L	35	Standard
	V	51	6482.3	2.1	**0.5523**	0.015	2.6	ug/L	1435	Standard
	Cr	52	11556.7	2.9	**0.3347**	0.014	4.0	ug/L	7635	Standard
	Cr	53	865.0	10.6	**0.4848**	0.072	14.9	ug/L	363	Standard
	Mn	55	123099.8	1.9	**15.2575**	0.218	1.4	ug/L	747	Standard
	Co	59	2101.5	1.5	**0.2401**	0.008	3.5	ug/L	234	Standard
	Ni	60	1463.7	4.5	**0.5572**	0.024	4.3	ug/L	215	Standard
	Cu	65	1004.0	2.5	**0.3664**	0.012	3.3	ug/L	104	Standard
	Zn	66	6132.9	2.2	**4.4370**	0.037	0.8	ug/L	363	Standard
>	**Ge**	72	568314.4	1.9				ug/L	495142	Standard
	As	75	135.4	21.1	**0.1651**	0.024	14.5	ug/L	-50	Standard
	Se	82	20.1	25.4	**0.0217**	0.043	197.6	ug/L	14	Standard
	Se-1	77	47.3	2.4	**-0.0575**	0.019	32.3	ug/L	42	Standard
>	**Ga**	71	615.0	5.6				mg/L	42	Standard
	Rb	85	5134.2	2.0				ug/L	17	Standard
	Y	89	506159.2	1.9				ug/L	417098	Standard
>	**Rh**	103	11.7	107.9				ug/L	13	Standard
	Mo	98	50.7	18.0	**0.0105**	0.003	28.3	ug/L	11	Standard
	Ag	107	125.7	10.6	**0.0035**	0.002	48.0	ug/L	68	Standard
	Cd	111	10.9	48.5	**-0.0035**	0.002	62.7	mg/L	2	Standard
	Cd	114	42.4	34.4	**0.0049**	0.002	48.6	ug/L	12	Standard
>	**In**	115	689821.8	0.4				ug/L	587441	Standard
	Sn	118	595.0	5.8	**-0.0080**	0.005	59.8	ug/L	497	Standard
	Sb	123	118.9	14.7	**0.0175**	0.003	18.8	ug/L	78	Standard
	Ba	135	10423.6	1.2	**4.3208**	0.035	0.8	ug/L	24	Standard
	Ce	140	31651.4	0.9				ug/L	25	Standard
>	**Tb**	159	1043112.5	1.4				ug/L	939701	Standard
	Ho	165	500.0	7.2				ug/L	10	Standard
	Tl	203	34.0	12.8	**0.0020**	0.000	23.2	ug/L	4	Standard
	Tl	205	20.0	66.1	**0.0027**	0.002	61.7	ug/L	2	Standard
	Pb	206	2645.2	1.4	**0.3784**	0.005	1.3	ug/L	240	Standard
	Pb	207	2160.5	2.6	**0.3433**	0.011	3.1	ug/L	198	Standard
	Pb	208	9496.8	0.4	**0.3858**	0.004	1.1	ug/L	883	Standard
	U	238	183.7	2.7	**0.0198**	0.001	4.5	ug/L	3	Standard
>	**Bi**	209	577076.9	1.0				ug/L	523984	Standard

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:38:07
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**15.9398**	27.600	173.2	mg/L	0	Standard
Mg	24	31.7	39.7	**-0.0310**	0.014	45.0	mg/L	37	Standard
K	39	75.0	17.6	**0.0236**	0.024	101.0	mg/L	38	Standard
Ca	43	45.0	44.4	**-0.1502**	1.087	724.0	mg/L	10	Standard
Fe	54	679.1	0.8	**0.3402**	0.010	2.9	mg/L	179	Standard
Fe	57	390.0	9.7	**0.0271**	0.102	376.6	mg/L	335	Standard
Sc-1	45	73444.3	1.5				mg/L	61029	Standard
Cl	35	27695.3	0.4				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	2046.8	15.0				ug/L	1547	Standard
P	31	20195.1	3.7				ug/L	15286	Standard
S	34	2511.9	7.3				ug/L	1572	Standard
Sr	88	68.3	42.9				ug/L	73	Standard
C	12	50.0	40.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	1042.0	9.5				mg/L	6	Standard
Ho-1	165	500.0	7.2				mg/L	10	Standard
Er	166	446.7	15.2				mg/L	13	Standard
I	127	1841.8	32.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		109.473	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.778	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138105
Report Date/Time: Wednesday, June 29, 2016 19:38:07
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		117.428
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.132
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138107
Sample Date/Time: Wednesday, June 29, 2016 19:39:01
Number of Replicates: 3
Autosampler Position: 402
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	84183.2	0.7				ug/L	74699	Standard
	Be	9	1125.0	5.6	**1.1418**	0.071	6.2	ug/L	8	Standard
	Al	27	16362118.1	1.0	**98.7783**	0.357	0.4	ug/L	1493	Standard
	Sc	45	80079.5	2.6				ug/L	61029	Standard
	Ti	47	5684.7	1.2	**13.0248**	0.178	1.4	ug/L	35	Standard
	V	51	193164.5	0.6	**22.2022**	0.395	1.8	ug/L	1435	Standard
	Cr	52	144926.2	1.3	**16.0993**	0.107	0.7	ug/L	7635	Standard
	Cr	53	17089.6	2.1	**15.8598**	0.165	1.0	ug/L	363	Standard
	Mn	55	4945150.4	1.2	**621.8545**	12.591	2.0	ug/L	747	Standard
	Co	59	79442.1	0.4	**10.4972**	0.101	1.0	ug/L	234	Standard
	Ni	60	48759.0	0.6	**22.6802**	0.156	0.7	ug/L	215	Standard
	Cu	65	44548.4	1.0	**20.8546**	0.085	0.4	ug/L	104	Standard
	Zn	66	116987.3	1.0	**93.3289**	1.249	1.3	ug/L	363	Standard
>	**Ge**	72	565487.8	1.3				ug/L	495142	Standard
	As	75	12786.6	1.5	**10.0557**	0.155	1.5	ug/L	-50	Standard
	Se	82	93.9	6.5	**0.6112**	0.058	9.6	ug/L	14	Standard
	Se-1	77	134.7	12.3	**0.9941**	0.220	22.1	ug/L	42	Standard
>	**Ga**	71	20750.9	1.1				mg/L	42	Standard
	Rb	85	229576.4	1.2				ug/L	17	Standard
	Y	89	760564.3	0.9				ug/L	417098	Standard
>	**Rh**	103	26.7	39.0				ug/L	13	Standard
	Mo	98	714.7	6.9	**0.2279**	0.017	7.3	ug/L	11	Standard
	Ag	107	706.7	2.8	**0.0816**	0.003	4.3	ug/L	68	Standard
	Cd	111	991.8	3.3	**0.4053**	0.015	3.7	mg/L	2	Standard
	Cd	114	2256.1	2.3	**0.3810**	0.005	1.2	ug/L	12	Standard
>	**In**	115	674434.0	1.0				ug/L	587441	Standard
	Sn	118	836.7	10.4	**0.0300**	0.012	41.1	ug/L	497	Standard
	Sb	123	542.0	6.0	**0.0969**	0.005	5.6	ug/L	78	Standard
	Ba	135	363437.5	1.5	**154.6607**	2.916	1.9	ug/L	24	Standard
	Ce	140	1323769.8	0.4				ug/L	25	Standard
>	**Tb**	159	1046974.9	0.8				ug/L	939701	Standard
	Ho	165	21420.1	0.8				ug/L	10	Standard
	Tl	203	1015.7	4.2	**0.1138**	0.005	4.5	ug/L	4	Standard
	Tl	205	906.7	2.2	**0.1191**	0.002	2.0	ug/L	2	Standard
	Pb	206	225485.9	0.9	**39.5300**	0.272	0.7	ug/L	240	Standard
	Pb	207	184080.5	0.2	**36.3178**	0.069	0.2	ug/L	198	Standard
	Pb	208	811320.1	0.8	**40.5199**	0.434	1.1	ug/L	883	Standard
	U	238	5280.9	2.0	**0.7154**	0.017	2.3	ug/L	3	Standard
>	**Bi**	209	558508.9	0.3				ug/L	523984	Standard

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:41:18
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	105.0	49.7	**0.0390**	0.054	139.2	mg/L	37	Standard
K	39	875.0	3.0	**1.3597**	0.017	1.3	mg/L	38	Standard
Ca	43	123.3	10.2	**3.4107**	0.543	15.9	mg/L	10	Standard
Fe	54	18540.3	2.4	**11.8486**	0.489	4.1	mg/L	179	Standard
Fe	57	5050.8	2.1	**12.0198**	0.614	5.1	mg/L	335	Standard
> **Sc-1**	45	80079.5	2.6				mg/L	61029	Standard
Cl	35	26211.3	2.3				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	4010.5	1.9				ug/L	1547	Standard
P	31	18821.7	2.8				ug/L	15286	Standard
S	34	2180.2	5.1				ug/L	1572	Standard
Sr	88	100.0	25.0				ug/L	73	Standard
C	12	166.7	30.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	70.0	28.6				mg/L	3	Standard
Dy	164	35314.8	2.4				mg/L	6	Standard
Ho-1	165	21420.1	0.8				mg/L	10	Standard
Er	166	18965.2	4.5				mg/L	13	Standard
I	127	21211.5	3.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		112.697	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.207	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:41:18
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	114.809
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	106.589
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:41:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138107
Sample Date/Time: Wednesday, June 29, 2016 19:42:13
Number of Replicates: 3
Autosampler Position: 403
Sample Description: 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79962.1	1.1				ug/L	74699	Standard
	Be	9	35.0	37.8	**0.0449**	0.015	32.4	ug/L	8	Standard
	Al	27	348986.4	2.3	**2.2050**	0.043	1.9	ug/L	1493	Standard
	Sc	45	69050.4	2.1				ug/L	61029	Standard
	Ti	47	147.7	4.0	**0.2190**	0.009	4.2	ug/L	35	Standard
	V	51	5537.0	5.1	**0.4222**	0.040	9.5	ug/L	1435	Standard
	Cr	52	11219.2	0.9	**0.2513**	0.015	5.8	ug/L	7635	Standard
	Cr	53	801.7	3.8	**0.4006**	0.025	6.2	ug/L	363	Standard
	Mn	55	101661.8	3.0	**12.1540**	0.524	4.3	ug/L	747	Standard
	Co	59	1813.1	5.5	**0.1941**	0.015	7.9	ug/L	234	Standard
	Ni	60	1251.7	0.2	**0.4392**	0.009	2.1	ug/L	215	Standard
	Cu	65	1098.7	4.3	**0.3937**	0.026	6.7	ug/L	104	Standard
	Zn	66	4663.4	1.9	**3.1415**	0.081	2.6	ug/L	363	Standard
>	**Ge**	72	587910.1	1.4				ug/L	495142	Standard
	As	75	208.3	5.8	**0.2162**	0.008	3.9	ug/L	-50	Standard
	Se	82	14.6	52.1	**-0.0264**	0.056	213.9	ug/L	14	Standard
	Se-1	77	58.3	15.6	**0.0509**	0.111	217.0	ug/L	42	Standard
>	**Ga**	71	483.3	11.4				mg/L	42	Standard
	Rb	85	4672.4	3.2				ug/L	17	Standard
	Y	89	494582.5	2.5				ug/L	417098	Standard
>	**Rh**	103	10.0	50.0				ug/L	13	Standard
	Mo	98	50.9	2.1	**0.0113**	0.001	5.2	ug/L	11	Standard
	Ag	107	137.0	5.3	**0.0059**	0.001	14.4	ug/L	68	Standard
	Cd	111	22.2	22.2	**0.0016**	0.002	123.1	mg/L	2	Standard
	Cd	114	55.9	25.6	**0.0077**	0.003	34.0	ug/L	12	Standard
>	**In**	115	657881.3	1.7				ug/L	587441	Standard
	Sn	118	586.7	15.6	**-0.0052**	0.012	239.2	ug/L	497	Standard
	Sb	123	130.5	4.0	**0.0208**	0.001	5.8	ug/L	78	Standard
	Ba	135	7584.6	1.4	**3.2933**	0.054	1.6	ug/L	24	Standard
	Ce	140	27601.8	1.0				ug/L	25	Standard
>	**Tb**	159	1004279.0	0.4				ug/L	939701	Standard
	Ho	165	470.0	12.8				ug/L	10	Standard
	Tl	203	31.3	21.7	**0.0018**	0.001	42.5	ug/L	4	Standard
	Tl	205	38.3	39.8	**0.0052**	0.002	37.2	ug/L	2	Standard
	Pb	206	5039.5	2.9	**0.8074**	0.036	4.5	ug/L	240	Standard
	Pb	207	4082.6	3.5	**0.7312**	0.039	5.3	ug/L	198	Standard
	Pb	208	17367.6	1.6	**0.7884**	0.025	3.2	ug/L	883	Standard
	U	238	118.3	14.2	**0.0116**	0.002	21.2	ug/L	3	Standard
>	**Bi**	209	562827.5	1.3				ug/L	523984	Standard

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:44:30
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	41.7	6.9	**-0.0175**	0.004	24.7	mg/L	37	Standard
K	39	65.0	46.8	**0.0126**	0.058	464.1	mg/L	38	Standard
Ca	43	20.0	66.1	**-1.4118**	0.719	50.9	mg/L	10	Standard
Fe	54	548.6	17.9	**0.2728**	0.072	26.3	mg/L	179	Standard
Fe	57	318.3	8.7	**-0.1165**	0.103	88.6	mg/L	335	Standard
> **Sc-1**	45	69050.4	2.1				mg/L	61029	Standard
Cl	35	24683.4	2.1				ug/L	20195	Standard
Kr	83	3.0	33.3				ug/L	4	Standard
Br	81	2036.8	6.4				ug/L	1547	Standard
P	31	17868.9	0.9				ug/L	15286	Standard
S	34	1880.1	5.1				ug/L	1572	Standard
Sr	88	70.0	32.7				ug/L	73	Standard
C	12	63.3	32.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	851.3	21.1				mg/L	6	Standard
Ho-1	165	470.0	12.8				mg/L	10	Standard
Er	166	460.0	19.0				mg/L	13	Standard
I	127	2083.5	32.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		107.046	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		118.736	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138107
Report Date/Time: Wednesday, June 29, 2016 19:44:30
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.991
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.413
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138402
Sample Date/Time: Wednesday, June 29, 2016 19:45:24
Number of Replicates: 3
Autosampler Position: 404
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78829.6	2.0				ug/L	74699	Standard
	Be	9	476.7	4.8	**0.5211**	0.034	6.6	ug/L	8	Standard
	Al	27	790696.4	0.9	**5.0858**	0.055	1.1	ug/L	1493	Standard
	Sc	45	79546.6	1.7				ug/L	61029	Standard
	Ti	47	5657.7	0.8	**12.8152**	0.349	2.7	ug/L	35	Standard
	V	51	106854.5	1.6	**12.0511**	0.193	1.6	ug/L	1435	Standard
	Cr	52	263051.4	1.4	**29.7045**	0.729	2.5	ug/L	7635	Standard
	Cr	53	32229.3	4.0	**29.8443**	0.696	2.3	ug/L	363	Standard
	Mn	55	1259215.6	0.6	**156.4341**	3.728	2.4	ug/L	747	Standard
	Co	59	45647.0	0.6	**5.9475**	0.157	2.6	ug/L	234	Standard
	Ni	60	134444.3	1.6	**62.0317**	0.857	1.4	ug/L	215	Standard
	Cu	65	50137.5	0.5	**23.2180**	0.523	2.3	ug/L	104	Standard
	Zn	66	191528.7	0.9	**151.3328**	2.746	1.8	ug/L	363	Standard
>	**Ge**	72	572152.6	2.7				ug/L	495142	Standard
	As	75	5700.3	1.6	**4.4645**	0.048	1.1	ug/L	-50	Standard
	Se	82	330.0	4.2	**2.4632**	0.049	2.0	ug/L	14	Standard
	Se-1	77	314.7	3.0	**3.1071**	0.115	3.7	ug/L	42	Standard
>	**Ga**	71	3685.4	1.6				mg/L	42	Standard
	Rb	85	41053.9	1.7				ug/L	17	Standard
	Y	89	1388727.0	1.6				ug/L	417098	Standard
>	**Rh**	103	25.0	0.0				ug/L	13	Standard
	Mo	98	3897.2	2.0	**1.2782**	0.019	1.5	ug/L	11	Standard
	Ag	107	551.3	5.8	**0.0614**	0.003	5.7	ug/L	68	Standard
	Cd	111	1240.5	0.7	**0.5131**	0.007	1.4	mg/L	2	Standard
	Cd	114	2915.5	4.4	**0.4970**	0.020	4.1	ug/L	12	Standard
>	**In**	115	669008.5	1.2				ug/L	587441	Standard
	Sn	118	676.7	8.2	**0.0070**	0.008	114.7	ug/L	497	Standard
	Sb	123	186.6	3.8	**0.0309**	0.001	3.8	ug/L	78	Standard
	Ba	135	221251.7	0.3	**94.9139**	1.345	1.4	ug/L	24	Standard
	Ce	140	471655.7	0.9				ug/L	25	Standard
>	**Tb**	159	1067869.0	0.9				ug/L	939701	Standard
	Ho	165	47698.3	1.5				ug/L	10	Standard
	Tl	203	1397.7	2.5	**0.1543**	0.005	3.4	ug/L	4	Standard
	Tl	205	1176.7	6.0	**0.1515**	0.008	5.0	ug/L	2	Standard
	Pb	206	21429.5	1.2	**3.6210**	0.069	1.9	ug/L	240	Standard
	Pb	207	15728.2	0.6	**2.9798**	0.021	0.7	ug/L	198	Standard
	Pb	208	70387.8	1.2	**3.3811**	0.084	2.5	ug/L	883	Standard
	U	238	18087.1	1.2	**2.4141**	0.054	2.2	ug/L	3	Standard
>	**Bi**	209	569388.1	1.3				ug/L	523984	Standard

Sample ID: L1606138402
Report Date/Time: Wednesday, June 29, 2016 19:47:41
Page 1

Approved: July 04, 2016

Na	23	5.0	173.2	**43.9188**	76.061	173.2	mg/L	0	Standard
Mg	24	210.0	6.3	**0.1429**	0.013	9.0	mg/L	37	Standard
K	39	275.0	8.3	**0.3521**	0.037	10.4	mg/L	38	Standard
Ca	43	148.3	19.8	**4.6742**	1.471	31.5	mg/L	10	Standard
Fe	54	12478.7	2.5	**7.9817**	0.313	3.9	mg/L	179	Standard
Fe	57	3430.4	4.1	**7.8688**	0.221	2.8	mg/L	335	Standard
> **Sc-1**	45	79546.6	1.7				mg/L	61029	Standard
Cl	35	26111.8	0.7				ug/L	20195	Standard
Kr	83	1.3	43.3				ug/L	4	Standard
Br	81	2043.5	3.5				ug/L	1547	Standard
P	31	19195.5	2.0				ug/L	15286	Standard
S	34	2368.5	5.4				ug/L	1572	Standard
Sr	88	153.3	13.2				ug/L	73	Standard
C	12	70.0	14.3				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	140.0	18.9				mg/L	3	Standard
Dy	164	65719.4	1.8				mg/L	6	Standard
Ho-1	165	47698.3	1.5				mg/L	10	Standard
Er	166	45979.7	1.2				mg/L	13	Standard
I	127	5242.6	1.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		105.530	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.553	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138402
Report Date/Time: Wednesday, June 29, 2016 19:47:41
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.885
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.665
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138404
Sample Date/Time: Wednesday, June 29, 2016 19:48:35
Number of Replicates: 3
Autosampler Position: 405
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	80955.8	1.9				ug/L	74699	Standard
	Be	9	880.0	13.6	**0.9302**	0.126	13.6	ug/L	8	Standard
	Al	27	4591606.5	1.6	**28.8172**	0.235	0.8	ug/L	1493	Standard
	Sc	45	81773.4	1.2				ug/L	61029	Standard
	Ti	47	8450.7	1.8	**19.2499**	0.245	1.3	ug/L	35	Standard
	V	51	185955.1	1.6	**21.1827**	0.308	1.5	ug/L	1435	Standard
	Cr	52	431888.4	0.7	**49.5778**	0.407	0.8	ug/L	7635	Standard
	Cr	53	54036.2	1.9	**50.4433**	1.377	2.7	ug/L	363	Standard
	Mn	55	1979621.7	0.4	**246.7487**	3.142	1.3	ug/L	747	Standard
	Co	59	89211.3	2.3	**11.6929**	0.262	2.2	ug/L	234	Standard
	Ni	60	241828.7	1.1	**112.0223**	0.265	0.2	ug/L	215	Standard
	Cu	65	79842.9	1.5	**37.1424**	0.248	0.7	ug/L	104	Standard
	Zn	66	363879.1	1.7	**288.7674**	1.874	0.6	ug/L	363	Standard
>	**Ge**	72	570278.1	1.3				ug/L	495142	Standard
	As	75	11266.0	1.4	**8.7923**	0.037	0.4	ug/L	-50	Standard
	Se	82	562.8	1.7	**4.3149**	0.122	2.8	ug/L	14	Standard
	Se-1	77	468.7	10.3	**4.9506**	0.587	11.9	ug/L	42	Standard
>	**Ga**	71	6996.6	2.8				mg/L	42	Standard
	Rb	85	75201.3	3.4				ug/L	17	Standard
	Y	89	1637404.7	1.2				ug/L	417098	Standard
>	**Rh**	103	40.0	12.5				ug/L	13	Standard
	Mo	98	6065.7	2.4	**2.0076**	0.060	3.0	ug/L	11	Standard
	Ag	107	921.4	1.6	**0.1122**	0.003	2.7	ug/L	68	Standard
	Cd	111	1933.3	0.9	**0.8101**	0.015	1.9	mg/L	2	Standard
	Cd	114	4671.7	0.8	**0.8037**	0.006	0.7	ug/L	12	Standard
>	**In**	115	664089.4	1.0				ug/L	587441	Standard
	Sn	118	708.3	7.1	**0.0125**	0.007	53.2	ug/L	497	Standard
	Sb	123	171.2	9.0	**0.0282**	0.003	10.1	ug/L	78	Standard
	Ba	135	217409.8	1.0	**93.9488**	0.914	1.0	ug/L	24	Standard
	Ce	140	744689.2	0.9				ug/L	25	Standard
>	**Tb**	159	1077211.2	1.7				ug/L	939701	Standard
	Ho	165	62662.7	1.1				ug/L	10	Standard
	Tl	203	2865.3	1.3	**0.3204**	0.004	1.2	ug/L	4	Standard
	Tl	205	2535.2	5.2	**0.3286**	0.016	4.9	ug/L	2	Standard
	Pb	206	40598.7	0.9	**6.9743**	0.127	1.8	ug/L	240	Standard
	Pb	207	29975.3	2.2	**5.7854**	0.186	3.2	ug/L	198	Standard
	Pb	208	132785.8	0.6	**6.4904**	0.091	1.4	ug/L	883	Standard
	U	238	27584.4	0.2	**3.7106**	0.059	1.6	ug/L	3	Standard
>	**Bi**	209	565317.9	1.5				ug/L	523984	Standard

Sample ID: L1606138404
Report Date/Time: Wednesday, June 29, 2016 19:50:52
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**14.1527**	24.505	173.1	mg/L	0	Standard
Mg	24	243.3	11.7	**0.1691**	0.025	14.5	mg/L	37	Standard
K	39	461.7	5.5	**0.6478**	0.049	7.5	mg/L	38	Standard
Ca	43	205.0	4.9	**7.1444**	0.559	7.8	mg/L	10	Standard
Fe	54	23854.6	1.3	**14.9557**	0.059	0.4	mg/L	179	Standard
Fe	57	6316.3	1.9	**14.9472**	0.123	0.8	mg/L	335	Standard
> **Sc-1**	45	81773.4	1.2				mg/L	61029	Standard
Cl	35	26395.0	1.0				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	8435.7	8.5				ug/L	1547	Standard
P	31	20819.3	1.9				ug/L	15286	Standard
S	34	2495.2	1.6				ug/L	1572	Standard
Sr	88	186.7	12.1				ug/L	73	Standard
C	12	26.7	21.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	520.0	8.4				mg/L	3	Standard
Dy	164	85733.4	1.4				mg/L	6	Standard
Ho-1	165	62662.7	1.1				mg/L	10	Standard
Er	166	61347.3	2.7				mg/L	13	Standard
I	127	12962.2	2.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		108.376	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.175	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138404
Report Date/Time: Wednesday, June 29, 2016 19:50:52
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.048
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.888
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ni 60 Upper, S, EEE	Ni	60	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606138404
Report Date/Time: Wednesday, June 29, 2016 19:50:52
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 19:51:49
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	79632.1	2.5				ug/L	74699	Standard
	Be	9	53037.7	1.9	**56.5448**	1.909	3.4	ug/L	8	Standard
	Al	27	8288547.2	1.5	**52.9043**	0.899	1.7	ug/L	1493	Standard
	Sc	45	68251.7	0.4				ug/L	61029	Standard
	Ti	47	44871.7	1.2	**102.9405**	2.024	2.0	ug/L	35	Standard
	V	51	450952.2	1.7	**51.7787**	1.350	2.6	ug/L	1435	Standard
	Cr	52	441674.6	0.8	**50.8503**	0.824	1.6	ug/L	7635	Standard
	Cr	53	54163.4	3.4	**50.6767**	1.716	3.4	ug/L	363	Standard
	Mn	55	408648.7	3.2	**50.9466**	2.025	4.0	ug/L	747	Standard
	Co	59	404996.9	2.1	**53.3495**	1.595	3.0	ug/L	234	Standard
	Ni	60	109209.9	1.2	**50.6487**	1.081	2.1	ug/L	215	Standard
	Cu	65	108068.7	1.0	**50.4372**	0.990	2.0	ug/L	104	Standard
	Zn	66	63914.2	1.2	**50.4711**	0.538	1.1	ug/L	363	Standard
>	Ge	72	568913.3	1.0				ug/L	495142	Standard
	As	75	64596.5	1.7	**50.2573**	1.198	2.4	ug/L	-50	Standard
	Se	82	6434.6	2.8	**50.8949**	1.678	3.3	ug/L	14	Standard
	Se-1	77	4404.6	0.9	**51.8699**	0.916	1.8	ug/L	42	Standard
>	Ga	71	81.7	19.7				mg/L	42	Standard
	Rb	85	868.4	8.5				ug/L	17	Standard
	Y	89	484608.1	0.7				ug/L	417098	Standard
>	Rh	103	25.0	40.0				ug/L	13	Standard
	Mo	98	300990.3	2.1	**101.3882**	1.976	1.9	ug/L	11	Standard
	Ag	107	359716.1	0.6	**49.5455**	0.289	0.6	ug/L	68	Standard
	Cd	111	115140.5	0.5	**49.4382**	0.205	0.4	mg/L	2	Standard
	Cd	114	277323.3	2.2	**48.5511**	1.074	2.2	ug/L	12	Standard
>	In	115	654244.8	0.1				ug/L	587441	Standard
	Sn	118	326410.3	0.9	**50.0768**	0.438	0.9	ug/L	497	Standard
	Sb	123	257734.5	0.6	**49.5626**	0.244	0.5	ug/L	78	Standard
	Ba	135	107937.1	0.2	**47.3350**	0.075	0.2	ug/L	24	Standard
	Ce	140	156.7	33.8				ug/L	25	Standard
>	Tb	159	993573.1	1.0				ug/L	939701	Standard
	Ho	165	18.3	41.7				ug/L	10	Standard
	Tl	203	442343.3	0.6	**51.3703**	0.497	1.0	ug/L	4	Standard
	Tl	205	397679.6	3.2	**53.2177**	1.929	3.6	ug/L	2	Standard
	Pb	206	282572.7	0.8	**50.5712**	0.781	1.5	ug/L	240	Standard
	Pb	207	252130.0	1.0	**50.7898**	0.969	1.9	ug/L	198	Standard
	Pb	208	1034503.8	1.3	**52.7468**	1.180	2.2	ug/L	883	Standard
	U	238	372072.1	1.5	**51.7486**	1.271	2.5	ug/L	3	Standard
>	Bi	209	547364.2	0.9				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:54:05
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.1101**	29.627	173.2	mg/L	0	Standard
Mg	24	4437.3	1.9	**5.0568**	0.105	2.1	mg/L	37	Standard
K	39	2166.8	4.9	**4.1677**	0.193	4.6	mg/L	38	Standard
Ca	43	105.0	40.7	**3.4097**	2.426	71.2	mg/L	10	Standard
Fe	54	6964.2	0.4	**5.1415**	0.037	0.7	mg/L	179	Standard
Fe	57	2053.5	5.5	**5.1659**	0.318	6.1	mg/L	335	Standard
> **Sc-1**	45	68251.7	0.4				mg/L	61029	Standard
Cl	35	27215.8	2.7				ug/L	20195	Standard
Kr	83	2.0	86.6				ug/L	4	Standard
Br	81	1920.1	6.9				ug/L	1547	Standard
P	31	20039.9	1.0				ug/L	15286	Standard
S	34	2546.9	7.4				ug/L	1572	Standard
Sr	88	71.7	29.0				ug/L	73	Standard
C	12	70.0	24.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	11.7	100.8				mg/L	6	Standard
Ho-1	165	18.3	41.7				mg/L	10	Standard
Er	166	33.3	34.6				mg/L	13	Standard
I	127	1696.8	22.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	113.090		
Al	27	105.809		
Sc	45			
Ti	47	102.941		
V	51	103.557		
Cr	52	101.701		
Cr	53			
Mn	55	101.893		
Co	59	106.699		
Ni	60	101.297		
Cu	65	100.874		
Zn	66	100.942		
> Ge	72		114.899	
As	75	100.515		
Se	82	101.790		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:54:05
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	101.388		
Ag	107	99.091		
Cd	111	98.876		
Cd	114			
In	115		111.372	
Sn	118	100.154		
Sb	123	99.125		
Ba	135	94.670		
Ce	140			
Tb	159			
Ho	165			
Tl	203	102.741		
Tl	205			
Pb	206			
Pb	207			
Pb	208	105.494		
U	238	103.497		
Bi	209		104.462	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 19:54:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 19:55:00
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78377.2	0.9				ug/L	74699	Standard
	Be	9	16.7	17.3	**0.0256**	0.003	11.6	ug/L	8	Standard
	Al	27	3955.5	8.5	**0.0123**	0.002	16.7	ug/L	1493	Standard
	Sc	45	70123.6	2.1				ug/L	61029	Standard
	Ti	47	44.3	19.2	**-0.0067**	0.022	331.3	ug/L	35	Standard
	V	51	1494.9	5.1	**-0.0231**	0.005	19.5	ug/L	1435	Standard
	Cr	52	7869.7	1.7	**-0.0976**	0.038	39.0	ug/L	7635	Standard
	Cr	53	438.3	4.0	**0.0839**	0.027	31.8	ug/L	363	Standard
	Mn	55	1323.4	11.3	**0.0175**	0.017	96.4	ug/L	747	Standard
	Co	59	292.0	15.1	**0.0012**	0.005	416.2	ug/L	234	Standard
	Ni	60	231.7	5.0	**-0.0160**	0.006	40.4	ug/L	215	Standard
	Cu	65	142.7	7.1	**-0.0369**	0.003	9.0	ug/L	104	Standard
	Zn	66	477.7	7.4	**-0.0736**	0.020	26.8	ug/L	363	Standard
>	**Ge**	72	568079.4	2.5				ug/L	495142	Standard
	As	75	-36.3	34.6	**0.0313**	0.010	31.8	ug/L	-50	Standard
	Se	82	17.1	26.2	**-0.0026**	0.033	1282.0	ug/L	14	Standard
	Se-1	77	54.3	9.3	**0.0268**	0.068	253.5	ug/L	42	Standard
>	**Ga**	71	40.0	12.5				mg/L	42	Standard
	Rb	85	38.3	19.9				ug/L	17	Standard
	Y	89	483139.1	0.9				ug/L	417098	Standard
>	**Rh**	103	11.7	99.0				ug/L	13	Standard
	Mo	98	291.7	36.0	**0.0908**	0.037	40.3	ug/L	11	Standard
	Ag	107	161.3	21.3	**0.0089**	0.005	57.1	ug/L	68	Standard
	Cd	111	21.8	28.2	**0.0012**	0.002	201.3	mg/L	2	Standard
	Cd	114	48.0	10.4	**0.0061**	0.001	15.7	ug/L	12	Standard
>	**In**	115	669206.2	1.8				ug/L	587441	Standard
	Sn	118	881.7	18.2	**0.0375**	0.022	58.1	ug/L	497	Standard
	Sb	123	447.3	19.7	**0.0797**	0.015	18.9	ug/L	78	Standard
	Ba	135	111.0	14.0	**0.0316**	0.007	21.6	ug/L	24	Standard
	Ce	140	185.0	42.5				ug/L	25	Standard
>	**Tb**	159	1027080.2	0.9				ug/L	939701	Standard
	Ho	165	28.3	10.2				ug/L	10	Standard
	Tl	203	102.0	27.3	**0.0095**	0.003	32.1	ug/L	4	Standard
	Tl	205	76.7	13.6	**0.0100**	0.001	13.1	ug/L	2	Standard
	Pb	206	326.3	10.0	**-0.0155**	0.005	35.3	ug/L	240	Standard
	Pb	207	292.3	4.6	**-0.0139**	0.002	17.4	ug/L	198	Standard
	Pb	208	1226.7	4.7	**-0.0145**	0.003	18.0	ug/L	883	Standard
	U	238	78.7	31.6	**0.0060**	0.003	54.4	ug/L	3	Standard
>	**Bi**	209	574845.6	0.5				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:57:17
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	34.4	**-0.0124**	0.019	155.4	mg/L	37	Standard
K	39	46.7	44.6	**-0.0238**	0.042	178.2	mg/L	38	Standard
Ca	43	26.7	57.3	**-1.0457**	0.868	83.0	mg/L	10	Standard
Fe	54	161.1	25.9	**-0.0190**	0.033	173.6	mg/L	179	Standard
Fe	57	251.7	19.3	**-0.3314**	0.129	38.8	mg/L	335	Standard
Sc-1	45	70123.6	2.1				mg/L	61029	Standard
Cl	35	27868.9	0.6				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	2040.1	1.8				ug/L	1547	Standard
P	31	20417.1	4.9				ug/L	15286	Standard
S	34	2425.2	3.5				ug/L	1572	Standard
Sr	88	45.0	40.1				ug/L	73	Standard
C	12	50.0	52.9				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	11.9	129.9				mg/L	6	Standard
Ho-1	165	28.3	10.2				mg/L	10	Standard
Er	166	30.0	57.7				mg/L	13	Standard
I	127	1460.1	42.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		114.731	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:57:17
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.919
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.707
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 19:57:17
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138406
Sample Date/Time: Wednesday, June 29, 2016 19:58:13
Number of Replicates: 3
Autosampler Position: 406
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77039.0	4.7				ug/L	74699	Standard
	Be	9	588.3	14.0	**0.6548**	0.074	11.3	ug/L	8	Standard
	Al	27	832772.7	0.7	**5.4883**	0.227	4.1	ug/L	1493	Standard
	Sc	45	75047.3	4.1				ug/L	61029	Standard
	Ti	47	6368.7	0.7	**14.6522**	0.311	2.1	ug/L	35	Standard
	V	51	122665.5	0.9	**14.0720**	0.287	2.0	ug/L	1435	Standard
	Cr	52	293324.5	0.5	**33.7422**	0.348	1.0	ug/L	7635	Standard
	Cr	53	35827.3	1.9	**33.7165**	0.271	0.8	ug/L	363	Standard
	Mn	55	1790039.2	2.5	**225.6481**	2.955	1.3	ug/L	747	Standard
	Co	59	59513.4	2.6	**7.8774**	0.097	1.2	ug/L	234	Standard
	Ni	60	162230.6	1.1	**75.9835**	0.734	1.0	ug/L	215	Standard
	Cu	65	58756.0	0.7	**27.6257**	0.286	1.0	ug/L	104	Standard
	Zn	66	213240.6	1.5	**171.0057**	1.221	0.7	ug/L	363	Standard
>	**Ge**	72	563753.3	1.5				ug/L	495142	Standard
	As	75	6354.9	1.3	**5.0439**	0.138	2.7	ug/L	-50	Standard
	Se	82	268.5	3.3	**2.0114**	0.102	5.1	ug/L	14	Standard
	Se-1	77	256.0	4.1	**2.4577**	0.168	6.8	ug/L	42	Standard
>	**Ga**	71	4010.5	2.8				mg/L	42	Standard
	Rb	85	39651.9	3.3				ug/L	17	Standard
	Y	89	1622063.0	2.1				ug/L	417098	Standard
>	**Rh**	103	33.3	8.7				ug/L	13	Standard
	Mo	98	4372.5	0.7	**1.4952**	0.034	2.3	ug/L	11	Standard
	Ag	107	608.3	2.2	**0.0725**	0.003	3.7	ug/L	68	Standard
	Cd	111	1369.4	3.4	**0.5916**	0.033	5.6	mg/L	2	Standard
	Cd	114	3220.4	7.1	**0.5731**	0.054	9.4	ug/L	12	Standard
>	**In**	115	642314.3	2.8				ug/L	587441	Standard
	Sn	118	866.7	3.3	**0.0411**	0.007	16.9	ug/L	497	Standard
	Sb	123	475.1	9.4	**0.0890**	0.010	11.7	ug/L	78	Standard
	Ba	135	210364.7	1.0	**94.0146**	1.722	1.8	ug/L	24	Standard
	Ce	140	620778.1	1.3				ug/L	25	Standard
>	**Tb**	159	1035282.0	2.1				ug/L	939701	Standard
	Ho	165	58499.4	0.8				ug/L	10	Standard
	Tl	203	1670.8	3.4	**0.1930**	0.007	3.5	ug/L	4	Standard
	Tl	205	1583.4	5.7	**0.2129**	0.013	6.3	ug/L	2	Standard
	Pb	206	23886.1	1.1	**4.2252**	0.039	0.9	ug/L	240	Standard
	Pb	207	17508.1	0.9	**3.4746**	0.041	1.2	ug/L	198	Standard
	Pb	208	79776.0	1.1	**4.0140**	0.064	1.6	ug/L	883	Standard
	U	238	22400.5	0.6	**3.1225**	0.026	0.8	ug/L	3	Standard
>	**Bi**	209	545342.1	0.8				ug/L	523984	Standard

Sample ID: L1606138406
Report Date/Time: Wednesday, June 29, 2016 20:00:30
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	190.0	5.3	**0.1344**	0.009	6.3	mg/L	37	Standard
K	39	241.7	12.5	**0.3203**	0.051	16.0	mg/L	38	Standard
Ca	43	168.3	3.4	**6.1267**	0.214	3.5	mg/L	10	Standard
Fe	54	12517.1	3.0	**8.4992**	0.364	4.3	mg/L	179	Standard
Fe	57	3428.7	4.4	**8.4118**	0.531	6.3	mg/L	335	Standard
> **Sc-1**	45	75047.3	4.1				mg/L	61029	Standard
Cl	35	26434.4	2.8				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2270.2	3.9				ug/L	1547	Standard
P	31	19440.8	0.5				ug/L	15286	Standard
S	34	2256.8	6.7				ug/L	1572	Standard
Sr	88	170.0	33.4				ug/L	73	Standard
C	12	76.7	54.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	300.0	18.6				mg/L	3	Standard
Dy	164	81135.2	0.1				mg/L	6	Standard
Ho-1	165	58499.4	0.8				mg/L	10	Standard
Er	166	56983.6	0.5				mg/L	13	Standard
I	127	5462.7	4.2				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.133	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.857	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138406
Report Date/Time: Wednesday, June 29, 2016 20:00:30
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	109.341
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	104.076
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138408
Sample Date/Time: Wednesday, June 29, 2016 20:01:24
Number of Replicates: 3
Autosampler Position: 407
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79667.3	1.9				ug/L	74699	Standard
	Be	9	540.0	10.4	**0.5836**	0.070	11.9	ug/L	8	Standard
	Al	27	892686.6	1.0	**5.6842**	0.164	2.9	ug/L	1493	Standard
	Sc	45	78111.0	3.7				ug/L	61029	Standard
	Ti	47	6171.3	0.9	**13.8487**	0.141	1.0	ug/L	35	Standard
	V	51	131186.8	0.8	**14.6946**	0.149	1.0	ug/L	1435	Standard
	Cr	52	291322.4	2.4	**32.6724**	0.743	2.3	ug/L	7635	Standard
	Cr	53	35120.7	4.3	**32.2467**	1.471	4.6	ug/L	363	Standard
	Mn	55	2233691.8	2.2	**274.8758**	6.127	2.2	ug/L	747	Standard
	Co	59	65245.2	1.7	**8.4325**	0.147	1.7	ug/L	234	Standard
	Ni	60	190260.1	2.2	**86.9887**	2.023	2.3	ug/L	215	Standard
	Cu	65	66483.6	1.3	**30.5171**	0.399	1.3	ug/L	104	Standard
	Zn	66	187493.0	1.2	**146.6893**	2.307	1.6	ug/L	363	Standard
>	**Ge**	72	577612.7	0.6				ug/L	495142	Standard
	As	75	6445.8	1.9	**4.9929**	0.116	2.3	ug/L	-50	Standard
	Se	82	222.3	2.5	**1.5982**	0.037	2.3	ug/L	14	Standard
	Se-1	77	243.0	1.1	**2.2298**	0.039	1.8	ug/L	42	Standard
>	**Ga**	71	4360.6	1.0				mg/L	42	Standard
	Rb	85	41170.9	1.0				ug/L	17	Standard
	Y	89	1636669.5	0.9				ug/L	417098	Standard
>	**Rh**	103	33.3	22.9				ug/L	13	Standard
	Mo	98	4750.5	1.5	**1.5922**	0.009	0.6	ug/L	11	Standard
	Ag	107	651.7	5.1	**0.0767**	0.003	4.5	ug/L	68	Standard
	Cd	111	1476.5	1.8	**0.6252**	0.009	1.4	mg/L	2	Standard
	Cd	114	3537.9	6.1	**0.6161**	0.029	4.7	ug/L	12	Standard
>	**In**	115	655238.7	1.5				ug/L	587441	Standard
	Sn	118	965.0	3.4	**0.0534**	0.004	7.7	ug/L	497	Standard
	Sb	123	268.7	11.3	**0.0474**	0.005	10.8	ug/L	78	Standard
	Ba	135	196413.8	1.6	**86.0176**	0.137	0.2	ug/L	24	Standard
	Ce	140	596487.3	1.9				ug/L	25	Standard
>	**Tb**	159	1073225.5	1.0				ug/L	939701	Standard
	Ho	165	61178.3	1.7				ug/L	10	Standard
	Tl	203	2456.5	1.8	**0.2718**	0.001	0.3	ug/L	4	Standard
	Tl	205	2150.2	4.8	**0.2763**	0.017	6.1	ug/L	2	Standard
	Pb	206	24200.3	1.9	**4.0880**	0.010	0.3	ug/L	240	Standard
	Pb	207	17656.3	0.7	**3.3464**	0.078	2.3	ug/L	198	Standard
	Pb	208	79187.0	3.1	**3.8025**	0.046	1.2	ug/L	883	Standard
	U	238	22940.4	1.4	**3.0557**	0.038	1.2	ug/L	3	Standard
>	**Bi**	209	570736.3	2.0				ug/L	523984	Standard

Sample ID: L1606138408
Report Date/Time: Wednesday, June 29, 2016 20:03:41
Page 1

Approved: July 04, 2016

Na	23	3.3	173.2	**28.7877**	49.853	173.2	mg/L	0	Standard
Mg	24	161.7	15.3	**0.0975**	0.019	19.1	mg/L	37	Standard
K	39	273.3	12.8	**0.3583**	0.064	17.7	mg/L	38	Standard
Ca	43	191.7	9.9	**6.9329**	0.825	11.9	mg/L	10	Standard
Fe	54	12889.7	4.6	**8.4015**	0.298	3.5	mg/L	179	Standard
Fe	57	3420.4	5.7	**8.0292**	0.829	10.3	mg/L	335	Standard
> **Sc-1**	45	78111.0	3.7				mg/L	61029	Standard
Cl	35	25108.8	1.1				ug/L	20195	Standard
Kr	83	1.0	173.2				ug/L	4	Standard
Br	81	1990.1	4.1				ug/L	1547	Standard
P	31	18960.2	2.1				ug/L	15286	Standard
S	34	2216.8	7.1				ug/L	1572	Standard
Sr	88	153.3	28.5				ug/L	73	Standard
C	12	110.0	9.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	313.3	12.9				mg/L	3	Standard
Dy	164	83986.8	2.5				mg/L	6	Standard
Ho-1	165	61178.3	1.7				mg/L	10	Standard
Er	166	57776.6	0.9				mg/L	13	Standard
I	127	4875.8	1.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		106.651	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		116.656	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138408
Report Date/Time: Wednesday, June 29, 2016 20:03:41
Page 2

Approved: July 04, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	111.541
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	108.922
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606138408
Report Date/Time: Wednesday, June 29, 2016 20:03:41
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138410
Sample Date/Time: Wednesday, June 29, 2016 20:04:36
Number of Replicates: 3
Autosampler Position: 408
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	82222.6	2.1				ug/L	74699	Standard
	Be	9	466.7	0.6	**0.4893**	0.010	2.0	ug/L	8	Standard
	Al	27	798979.6	1.9	**4.9264**	0.080	1.6	ug/L	1493	Standard
	Sc	45	79640.5	1.7				ug/L	61029	Standard
	Ti	47	5792.4	2.5	**13.1770**	0.593	4.5	ug/L	35	Standard
	V	51	95139.7	1.4	**10.7518**	0.246	2.3	ug/L	1435	Standard
	Cr	52	279542.5	2.0	**31.7533**	0.664	2.1	ug/L	7635	Standard
	Cr	53	34038.2	1.3	**31.6794**	0.880	2.8	ug/L	363	Standard
	Mn	55	1053790.0	0.3	**131.3975**	3.080	2.3	ug/L	747	Standard
	Co	59	35027.1	0.6	**4.5723**	0.080	1.8	ug/L	234	Standard
	Ni	60	156240.8	1.2	**72.3985**	1.638	2.3	ug/L	215	Standard
	Cu	65	48550.0	0.3	**22.5657**	0.405	1.8	ug/L	104	Standard
	Zn	66	257712.2	0.3	**204.5897**	4.503	2.2	ug/L	363	Standard
>	**Ge**	72	569889.2	2.0				ug/L	495142	Standard
	As	75	6132.6	2.6	**4.8186**	0.190	4.0	ug/L	-50	Standard
	Se	82	361.1	9.2	**2.7244**	0.314	11.5	ug/L	14	Standard
	Se-1	77	331.7	7.4	**3.3212**	0.233	7.0	ug/L	42	Standard
>	**Ga**	71	4152.2	2.1				mg/L	42	Standard
	Rb	85	52623.0	3.5				ug/L	17	Standard
	Y	89	1425089.1	0.3				ug/L	417098	Standard
>	**Rh**	103	33.3	8.7				ug/L	13	Standard
	Mo	98	3927.9	0.3	**1.2648**	0.018	1.4	ug/L	11	Standard
	Ag	107	467.0	2.9	**0.0489**	0.003	5.2	ug/L	68	Standard
	Cd	111	1169.0	3.4	**0.4740**	0.011	2.3	mg/L	2	Standard
	Cd	114	2891.7	1.0	**0.4840**	0.007	1.5	ug/L	12	Standard
>	**In**	115	681463.1	1.2				ug/L	587441	Standard
	Sn	118	755.0	5.0	**0.0168**	0.007	41.1	ug/L	497	Standard
	Sb	123	219.3	13.0	**0.0363**	0.005	13.2	ug/L	78	Standard
	Ba	135	73803.3	1.7	**31.0670**	0.270	0.9	ug/L	24	Standard
	Ce	140	554709.1	1.1				ug/L	25	Standard
>	**Tb**	159	1094015.2	1.4				ug/L	939701	Standard
	Ho	165	50373.6	1.6				ug/L	10	Standard
	Tl	203	1120.7	1.3	**0.1214**	0.002	2.0	ug/L	4	Standard
	Tl	205	941.7	7.2	**0.1194**	0.008	6.7	ug/L	2	Standard
	Pb	206	24028.0	1.2	**4.0020**	0.045	1.1	ug/L	240	Standard
	Pb	207	17504.4	0.3	**3.2701**	0.054	1.6	ug/L	198	Standard
	Pb	208	78967.6	0.8	**3.7402**	0.091	2.4	ug/L	883	Standard
	U	238	25248.7	0.7	**3.3178**	0.077	2.3	ug/L	3	Standard
>	**Bi**	209	578682.9	1.6				ug/L	523984	Standard

Sample ID: L1606138410
Report Date/Time: Wednesday, June 29, 2016 20:06:53
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	188.3	33.0	**0.1211**	0.061	50.1	mg/L	37	Standard
K	39	248.3	3.1	**0.3067**	0.020	6.4	mg/L	38	Standard
Ca	43	173.3	7.3	**5.8747**	0.743	12.6	mg/L	10	Standard
Fe	54	14056.9	1.9	**8.9942**	0.034	0.4	mg/L	179	Standard
Fe	57	3822.2	6.9	**8.8892**	0.835	9.4	mg/L	335	Standard
Sc-1	45	79640.5	1.7				mg/L	61029	Standard
Cl	35	25424.6	1.4				ug/L	20195	Standard
Kr	83	1.3	114.6				ug/L	4	Standard
Br	81	2053.5	6.1				ug/L	1547	Standard
P	31	20016.5	1.3				ug/L	15286	Standard
S	34	2301.8	5.1				ug/L	1572	Standard
Sr	88	141.7	32.0				ug/L	73	Standard
C	12	56.7	36.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	133.3	11.5				mg/L	3	Standard
Dy	164	72545.2	2.7				mg/L	6	Standard
Ho-1	165	50373.6	1.6				mg/L	10	Standard
Er	166	47160.0	0.9				mg/L	13	Standard
I	127	6799.9	2.8				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		110.072	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		115.096	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606138410
Report Date/Time: Wednesday, June 29, 2016 20:06:53
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.005
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.439
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606138412
Sample Date/Time: Wednesday, June 29, 2016 20:07:47
Number of Replicates: 3
Autosampler Position: 409
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81391.5	3.4				ug/L	74699	Standard
	Be	9	565.0	6.9	**0.5962**	0.020	3.4	ug/L	8	Standard
	Al	27	802320.4	2.6	**4.9984**	0.082	1.6	ug/L	1493	Standard
	Sc	45	77871.4	3.1				ug/L	61029	Standard
	Ti	47	8431.3	2.1	**19.3525**	0.252	1.3	ug/L	35	Standard
	V	51	136034.9	0.6	**15.5641**	0.178	1.1	ug/L	1435	Standard
	Cr	52	292890.6	0.7	**33.5584**	0.660	2.0	ug/L	7635	Standard
	Cr	53	35682.0	1.9	**33.4491**	0.599	1.8	ug/L	363	Standard
	Mn	55	2260782.0	0.5	**283.9966**	5.862	2.1	ug/L	747	Standard
	Co	59	82153.8	0.1	**10.8487**	0.184	1.7	ug/L	234	Standard
	Ni	60	172794.1	1.0	**80.6313**	1.536	1.9	ug/L	215	Standard
	Cu	65	63565.7	0.4	**29.7801**	0.432	1.5	ug/L	104	Standard
	Zn	66	251019.4	0.8	**200.6145**	2.591	1.3	ug/L	363	Standard
>	**Ge**	72	565970.9	1.7				ug/L	495142	Standard
	As	75	7636.9	2.3	**6.0250**	0.142	2.4	ug/L	-50	Standard
	Se	82	426.3	0.8	**3.2611**	0.084	2.6	ug/L	14	Standard
	Se-1	77	381.7	1.3	**3.9506**	0.115	2.9	ug/L	42	Standard
>	**Ga**	71	4992.5	2.6				mg/L	42	Standard
	Rb	85	46121.8	0.5				ug/L	17	Standard
	Y	89	1754842.6	1.7				ug/L	417098	Standard
>	**Rh**	103	43.3	35.3				ug/L	13	Standard
	Mo	98	5382.6	1.1	**1.7465**	0.044	2.5	ug/L	11	Standard
	Ag	107	637.3	4.9	**0.0719**	0.004	5.2	ug/L	68	Standard
	Cd	111	1430.5	6.2	**0.5853**	0.028	4.9	mg/L	2	Standard
	Cd	114	3407.8	7.3	**0.5746**	0.047	8.2	ug/L	12	Standard
>	**In**	115	677210.7	1.5				ug/L	587441	Standard
	Sn	118	780.0	4.5	**0.0212**	0.006	30.1	ug/L	497	Standard
	Sb	123	243.4	9.9	**0.0411**	0.005	12.6	ug/L	78	Standard
	Ba	135	199262.8	0.6	**84.4450**	1.189	1.4	ug/L	24	Standard
	Ce	140	678125.1	0.7				ug/L	25	Standard
>	**Tb**	159	1071624.9	1.1				ug/L	939701	Standard
	Ho	165	67051.4	0.5				ug/L	10	Standard
	Tl	203	1859.8	4.4	**0.2096**	0.007	3.5	ug/L	4	Standard
	Tl	205	1828.4	5.2	**0.2396**	0.011	4.7	ug/L	2	Standard
	Pb	206	33182.6	0.6	**5.7494**	0.057	1.0	ug/L	240	Standard
	Pb	207	24419.7	0.3	**4.7511**	0.035	0.7	ug/L	198	Standard
	Pb	208	111320.6	0.3	**5.4889**	0.050	0.9	ug/L	883	Standard
	U	238	25349.9	1.3	**3.4462**	0.014	0.4	ug/L	3	Standard
>	**Bi**	209	559224.1	1.0				ug/L	523984	Standard

Sample ID: L1606138412
Report Date/Time: Wednesday, June 29, 2016 20:10:04
Page 1

Approved: July 04, 2016

Na	23	3.3	86.6	**30.2574**	26.235	86.7	mg/L	0	Standard
Mg	24	168.3	9.1	**0.1052**	0.014	12.9	mg/L	37	Standard
K	39	280.0	6.4	**0.3709**	0.027	7.3	mg/L	38	Standard
Ca	43	176.7	27.2	**6.2736**	2.632	42.0	mg/L	10	Standard
Fe	54	15625.1	2.7	**10.2448**	0.046	0.4	mg/L	179	Standard
Fe	57	4357.3	3.7	**10.5343**	0.341	3.2	mg/L	335	Standard
> **Sc-1**	45	77871.4	3.1				mg/L	61029	Standard
Cl	35	26050.4	2.2				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2310.2	5.7				ug/L	1547	Standard
P	31	19966.5	2.1				ug/L	15286	Standard
S	34	2301.8	1.2				ug/L	1572	Standard
Sr	88	181.7	8.4				ug/L	73	Standard
C	12	60.0	66.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	466.7	15.8				mg/L	3	Standard
Dy	164	94635.2	0.7				mg/L	6	Standard
Ho-1	165	67051.4	0.5				mg/L	10	Standard
Er	166	62943.8	1.5				mg/L	13	Standard
I	127	6689.8	6.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		108.960	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.305	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606138412
Report Date/Time: Wednesday, June 29, 2016 20:10:04
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		115.281
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.725
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	

Sample ID: L1606138412
Report Date/Time: Wednesday, June 29, 2016 20:10:04
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 20:11:01
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	78720.7	2.9				ug/L	74699	Standard
	Be	9	53925.8	1.9	**58.1477**	1.206	2.1	ug/L	8	Standard
	Al	27	8443805.5	1.2	**54.5410**	1.847	3.4	ug/L	1493	Standard
	Sc	45	68074.4	3.5				ug/L	61029	Standard
	Ti	47	45146.2	1.1	**105.2966**	1.372	1.3	ug/L	35	Standard
	V	51	452407.6	0.9	**52.8112**	0.533	1.0	ug/L	1435	Standard
	Cr	52	439865.6	0.2	**51.4989**	0.494	1.0	ug/L	7635	Standard
	Cr	53	55131.8	1.5	**52.4610**	1.191	2.3	ug/L	363	Standard
	Mn	55	411766.6	1.7	**52.1886**	1.083	2.1	ug/L	747	Standard
	Co	59	411137.6	1.8	**55.0576**	1.090	2.0	ug/L	234	Standard
	Ni	60	108563.6	1.0	**51.1869**	0.633	1.2	ug/L	215	Standard
	Cu	65	108514.2	0.8	**51.4879**	0.273	0.5	ug/L	104	Standard
	Zn	66	63813.1	1.5	**51.2390**	0.656	1.3	ug/L	363	Standard
>	Ge	72	559563.9	0.9				ug/L	495142	Standard
	As	75	65025.3	0.7	**51.4299**	0.351	0.7	ug/L	-50	Standard
	Se	82	6607.2	0.8	**53.1327**	0.267	0.5	ug/L	14	Standard
	Se-1	77	4509.3	0.5	**54.0117**	0.383	0.7	ug/L	42	Standard
>	Ga	71	68.3	11.2				mg/L	42	Standard
	Rb	85	971.7	11.9				ug/L	17	Standard
	Y	89	484276.8	2.3				ug/L	417098	Standard
>	Rh	103	33.3	52.7				ug/L	13	Standard
	Mo	98	305352.5	1.7	**102.8803**	2.601	2.5	ug/L	11	Standard
	Ag	107	366186.0	1.2	**50.4411**	0.533	1.1	ug/L	68	Standard
	Cd	111	118185.8	1.5	**50.7502**	0.681	1.3	mg/L	2	Standard
	Cd	114	288421.2	1.0	**50.5001**	0.711	1.4	ug/L	12	Standard
>	In	115	654225.0	1.4				ug/L	587441	Standard
	Sn	118	325668.1	1.2	**49.9654**	0.174	0.3	ug/L	497	Standard
	Sb	123	260726.4	2.1	**50.1401**	0.797	1.6	ug/L	78	Standard
	Ba	135	109583.9	1.3	**48.0616**	0.591	1.2	ug/L	24	Standard
	Ce	140	191.7	20.9				ug/L	25	Standard
>	Tb	159	983539.2	1.4				ug/L	939701	Standard
	Ho	165	21.7	13.3				ug/L	10	Standard
	Tl	203	449254.6	0.7	**52.6820**	0.493	0.9	ug/L	4	Standard
	Tl	205	410287.0	1.8	**55.4353**	1.032	1.9	ug/L	2	Standard
	Pb	206	287176.7	1.5	**51.8952**	0.724	1.4	ug/L	240	Standard
	Pb	207	254226.6	1.1	**51.7082**	0.492	1.0	ug/L	198	Standard
	Pb	208	1062593.6	0.7	**54.7051**	0.438	0.8	ug/L	883	Standard
	U	238	383746.1	0.8	**53.8863**	0.314	0.6	ug/L	3	Standard
>	Bi	209	542059.3	0.3				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:13:18
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.9421**	31.068	173.2	mg/L	0	Standard
Mg	24	4572.4	3.8	**5.2291**	0.219	4.2	mg/L	37	Standard
K	39	2235.2	5.1	**4.3138**	0.121	2.8	mg/L	38	Standard
Ca	43	108.3	35.8	**3.6601**	2.380	65.0	mg/L	10	Standard
Fe	54	7126.2	2.9	**5.2851**	0.313	5.9	mg/L	179	Standard
Fe	57	2081.8	4.0	**5.2741**	0.321	6.1	mg/L	335	Standard
> **Sc-1**	45	68074.4	3.5				mg/L	61029	Standard
Cl	35	27726.7	0.9				ug/L	20195	Standard
Kr	83	0.7	86.6				ug/L	4	Standard
Br	81	1910.1	6.0				ug/L	1547	Standard
P	31	20502.2	2.0				ug/L	15286	Standard
S	34	2426.9	4.0				ug/L	1572	Standard
Sr	88	50.0	26.5				ug/L	73	Standard
C	12	73.3	20.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	15.9	37.3				mg/L	6	Standard
Ho-1	165	21.7	13.3				mg/L	10	Standard
Er	166	16.7	34.6				mg/L	13	Standard
I	127	1525.1	31.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	116.295		
Al	27	109.082		
Sc	45			
Ti	47	105.297		
V	51	105.622		
Cr	52	102.998		
Cr	53			
Mn	55	104.377		
Co	59	110.115		
Ni	60	102.374		
Cu	65	102.976		
Zn	66	102.478		
> Ge	72		113.011	
As	75	102.860		
Se	82	106.265		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:13:18
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	102.880		
Ag	107	100.882		
Cd	111	101.500		
Cd	114			
In	115		111.369	
Sn	118	99.931		
Sb	123	100.280		
Ba	135	96.123		
Ce	140			
Tb	159			
Ho	165			
Tl	203	105.364		
Tl	205			
Pb	206			
Pb	207			
Pb	208	109.410		
U	238	107.773		
Bi	209		103.450	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:13:18
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 20:14:12
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	79194.8	0.3				ug/L	74699	Standard
	Be	9	31.7	9.1	**0.0415**	0.003	7.4	ug/L	8	Standard
	Al	27	5331.0	39.7	**0.0208**	0.013	64.7	ug/L	1493	Standard
	Sc	45	70443.5	2.7				ug/L	61029	Standard
	Ti	47	45.0	15.6	**-0.0077**	0.016	206.1	ug/L	35	Standard
	V	51	1613.4	5.4	**-0.0133**	0.013	95.3	ug/L	1435	Standard
	Cr	52	7981.4	3.6	**-0.1047**	0.047	45.2	ug/L	7635	Standard
	Cr	53	418.3	25.3	**0.0570**	0.103	180.4	ug/L	363	Standard
	Mn	55	1741.1	30.3	**0.0658**	0.068	103.0	ug/L	747	Standard
	Co	59	378.7	17.0	**0.0117**	0.009	74.2	ug/L	234	Standard
	Ni	60	306.7	29.3	**0.0162**	0.043	265.9	ug/L	215	Standard
	Cu	65	193.0	22.7	**-0.0150**	0.021	143.0	ug/L	104	Standard
	Zn	66	576.0	19.1	**-0.0039**	0.093	2387.7	ug/L	363	Standard
>	**Ge**	72	580508.2	1.5				ug/L	495142	Standard
	As	75	-17.3	276.0	**0.0468**	0.036	77.6	ug/L	-50	Standard
	Se	82	19.7	47.6	**0.0162**	0.076	468.0	ug/L	14	Standard
	Se-1	77	62.3	8.8	**0.1062**	0.073	68.4	ug/L	42	Standard
>	**Ga**	71	43.3	24.0				mg/L	42	Standard
	Rb	85	53.3	27.1				ug/L	17	Standard
	Y	89	466004.6	0.7				ug/L	417098	Standard
>	**Rh**	103	18.3	41.7				ug/L	13	Standard
	Mo	98	326.5	17.1	**0.1045**	0.019	18.6	ug/L	11	Standard
	Ag	107	195.7	20.5	**0.0141**	0.005	38.2	ug/L	68	Standard
	Cd	111	49.8	49.3	**0.0134**	0.010	76.8	mg/L	2	Standard
	Cd	114	97.6	30.0	**0.0150**	0.005	33.6	ug/L	12	Standard
>	**In**	115	653349.0	0.9				ug/L	587441	Standard
	Sn	118	958.4	16.5	**0.0527**	0.023	43.7	ug/L	497	Standard
	Sb	123	468.2	24.3	**0.0858**	0.021	24.7	ug/L	78	Standard
	Ba	135	169.0	53.9	**0.0580**	0.039	67.5	ug/L	24	Standard
	Ce	140	383.3	72.7				ug/L	25	Standard
>	**Tb**	159	1008003.3	2.5				ug/L	939701	Standard
	Ho	165	33.3	56.8				ug/L	10	Standard
	Tl	203	138.0	27.1	**0.0136**	0.004	29.9	ug/L	4	Standard
	Tl	205	128.3	47.6	**0.0167**	0.008	47.6	ug/L	2	Standard
	Pb	206	364.7	12.5	**-0.0084**	0.008	93.5	ug/L	240	Standard
	Pb	207	321.7	11.0	**-0.0076**	0.006	84.7	ug/L	198	Standard
	Pb	208	1298.7	10.5	**-0.0103**	0.007	67.0	ug/L	883	Standard
	U	238	121.3	35.3	**0.0118**	0.006	48.9	ug/L	3	Standard
>	**Bi**	209	569222.6	0.7				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:16:29
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	31.7	50.8	**-0.0295**	0.019	63.3	mg/L	37	Standard
K	39	58.3	13.1	**-0.0021**	0.017	817.3	mg/L	38	Standard
Ca	43	26.7	28.6	**-1.0586**	0.442	41.7	mg/L	10	Standard
Fe	54	154.2	11.8	**-0.0252**	0.011	43.9	mg/L	179	Standard
Fe	57	258.3	4.9	**-0.3125**	0.056	18.0	mg/L	335	Standard
Sc-1	45	70443.5	2.7				mg/L	61029	Standard
Cl	35	27476.9	1.3				ug/L	20195	Standard
Kr	83	1.3	86.6				ug/L	4	Standard
Br	81	1946.8	3.9				ug/L	1547	Standard
P	31	20592.3	4.9				ug/L	15286	Standard
S	34	2348.5	2.8				ug/L	1572	Standard
Sr	88	75.0	46.7				ug/L	73	Standard
C	12	43.3	13.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	74.1	39.2				mg/L	6	Standard
Ho-1	165	33.3	56.8				mg/L	10	Standard
Er	166	53.3	71.0				mg/L	13	Standard
I	127	1293.4	30.4				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		117.241	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:16:29
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.219
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.633
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: PBW 43 WG573291-03
Sample Date/Time: Wednesday, June 29, 2016 20:17:26
Number of Replicates: 3
Autosampler Position: 410
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	78266.6	1.0				ug/L	74699	Standard
	Be	9	21.7	58.1	**0.0312**	0.014	44.4	ug/L	8	Standard
	Al	27	32901.2	14.4	**0.2003**	0.031	15.3	ug/L	1493	Standard
	Sc	45	68990.1	1.1				ug/L	61029	Standard
	Ti	47	59.7	25.6	**0.0271**	0.036	131.8	ug/L	35	Standard
	V	51	1740.6	13.9	**0.0036**	0.029	816.2	ug/L	1435	Standard
	Cr	52	10743.1	2.3	**0.2282**	0.033	14.5	ug/L	7635	Standard
	Cr	53	856.7	10.0	**0.4710**	0.086	18.2	ug/L	363	Standard
	Mn	55	4965.9	44.9	**0.4694**	0.280	59.7	ug/L	747	Standard
	Co	59	438.7	27.8	**0.0202**	0.016	81.0	ug/L	234	Standard
	Ni	60	464.7	17.9	**0.0906**	0.040	44.1	ug/L	215	Standard
	Cu	65	288.0	10.4	**0.0300**	0.015	49.7	ug/L	104	Standard
	Zn	66	2615.6	1.4	**1.6135**	0.041	2.5	ug/L	363	Standard
>	**Ge**	72	573366.1	0.8				ug/L	495142	Standard
	As	75	-37.9	81.3	**0.0303**	0.024	78.8	ug/L	-50	Standard
	Se	82	11.3	32.8	**-0.0488**	0.029	60.5	ug/L	14	Standard
	Se-1	77	61.0	9.1	**0.0990**	0.067	67.3	ug/L	42	Standard
>	**Ga**	71	31.7	9.1				mg/L	42	Standard
	Rb	85	100.0	26.5				ug/L	17	Standard
	Y	89	479627.1	2.2				ug/L	417098	Standard
>	**Rh**	103	1.7	173.2				ug/L	13	Standard
	Mo	98	155.9	47.8	**0.0477**	0.025	52.9	ug/L	11	Standard
	Ag	107	191.7	76.1	**0.0140**	0.020	145.7	ug/L	68	Standard
	Cd	111	21.5	58.8	**0.0014**	0.005	377.2	mg/L	2	Standard
	Cd	114	53.4	80.4	**0.0074**	0.008	102.3	ug/L	12	Standard
>	**In**	115	641242.2	1.6				ug/L	587441	Standard
	Sn	118	930.0	4.7	**0.0512**	0.008	15.1	ug/L	497	Standard
	Sb	123	286.5	19.6	**0.0520**	0.011	21.3	ug/L	78	Standard
	Ba	135	402.7	93.5	**0.1636**	0.167	102.4	ug/L	24	Standard
	Ce	140	951.7	120.0				ug/L	25	Standard
>	**Tb**	159	992014.1	1.0				ug/L	939701	Standard
	Ho	165	68.3	105.1				ug/L	10	Standard
	Tl	203	138.0	108.8	**0.0140**	0.017	122.9	ug/L	4	Standard
	Tl	205	136.7	67.7	**0.0182**	0.012	67.1	ug/L	2	Standard
	Pb	206	481.7	43.4	**0.0137**	0.037	272.5	ug/L	240	Standard
	Pb	207	395.0	43.8	**0.0084**	0.035	416.2	ug/L	198	Standard
	Pb	208	1692.0	26.2	**0.0108**	0.023	209.9	ug/L	883	Standard
	U	238	120.3	76.5	**0.0121**	0.013	104.7	ug/L	3	Standard
>	**Bi**	209	557465.1	0.7				ug/L	523984	Standard

Sample ID: PBW 43 WG573291-03
Report Date/Time: Wednesday, June 29, 2016 20:19:42
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	46.7	43.3	**-0.0120**	0.023	188.7	mg/L	37	Standard
K	39	41.7	73.3	**-0.0331**	0.059	178.0	mg/L	38	Standard
Ca	43	16.7	45.8	**-1.5897**	0.437	27.5	mg/L	10	Standard
Fe	54	190.8	15.9	**0.0047**	0.022	466.7	mg/L	179	Standard
Fe	57	281.7	18.5	**-0.2277**	0.154	67.5	mg/L	335	Standard
> **Sc-1**	45	68990.1	1.1				mg/L	61029	Standard
Cl	35	27241.8	2.0				ug/L	20195	Standard
Kr	83	3.3	34.6				ug/L	4	Standard
Br	81	1960.1	2.7				ug/L	1547	Standard
P	31	19042.0	2.0				ug/L	15286	Standard
S	34	2271.8	2.5				ug/L	1572	Standard
Sr	88	70.0	14.3				ug/L	73	Standard
C	12	93.3	12.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	57.4	72.5				mg/L	6	Standard
Ho-1	165	68.3	105.1				mg/L	10	Standard
Er	166	53.3	75.8				mg/L	13	Standard
I	127	1241.7	37.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		104.776	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.798	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBW 43 WG573291-03
Report Date/Time: Wednesday, June 29, 2016 20:19:42
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.159
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.390
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 43 WG573291-03
Report Date/Time: Wednesday, June 29, 2016 20:19:42
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: LCSW 43 WG573291-04
Sample Date/Time: Wednesday, June 29, 2016 20:20:37
Number of Replicates: 3
Autosampler Position: 411
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	79286.9	1.2				ug/L	74699	Standard
	Be	9	55518.2	1.7	**59.4178**	0.565	1.0	ug/L	8	Standard
	Al	27	18234.3	3.1	**0.1035**	0.005	4.4	ug/L	1493	Standard
	Sc	45	71925.5	3.7				ug/L	61029	Standard
	Ti	47	66.7	12.9	**0.0416**	0.023	54.2	ug/L	35	Standard
	V	51	464385.2	1.4	**52.3659**	1.151	2.2	ug/L	1435	Standard
	Cr	52	468709.1	0.6	**53.0500**	1.578	3.0	ug/L	7635	Standard
	Cr	53	57400.2	0.6	**52.7689**	1.471	2.8	ug/L	363	Standard
	Mn	55	425738.4	0.7	**52.1249**	1.035	2.0	ug/L	747	Standard
	Co	59	411335.9	1.2	**53.2163**	1.483	2.8	ug/L	234	Standard
	Ni	60	114540.7	1.0	**52.1842**	1.705	3.3	ug/L	215	Standard
	Cu	65	115424.0	0.4	**52.9151**	1.312	2.5	ug/L	104	Standard
	Zn	66	68318.0	0.2	**53.0163**	1.377	2.6	ug/L	363	Standard
>	Ge	72	579422.8	2.7				ug/L	495142	Standard
	As	75	63888.1	1.5	**48.8227**	1.412	2.9	ug/L	-50	Standard
	Se	82	6548.4	1.5	**50.8872**	2.152	4.2	ug/L	14	Standard
	Se-1	77	4570.0	2.1	**52.8570**	0.847	1.6	ug/L	42	Standard
>	Ga	71	68.3	27.7				mg/L	42	Standard
	Rb	85	96.7	7.9				ug/L	17	Standard
	Y	89	471340.1	1.5				ug/L	417098	Standard
>	Rh	103	33.3	52.7				ug/L	13	Standard
	Mo	98	1844.9	1.6	**0.6204**	0.021	3.4	ug/L	11	Standard
	Ag	107	386699.9	0.4	**53.6468**	1.108	2.1	ug/L	68	Standard
	Cd	111	120754.4	0.7	**52.2256**	1.320	2.5	mg/L	2	Standard
	Cd	114	290744.9	1.2	**51.2713**	1.454	2.8	ug/L	12	Standard
>	In	115	649724.4	1.8				ug/L	587441	Standard
	Sn	118	890.0	5.4	**0.0430**	0.007	15.1	ug/L	497	Standard
	Sb	123	259622.2	1.1	**50.2896**	1.419	2.8	ug/L	78	Standard
	Ba	135	113320.2	1.2	**50.0586**	1.415	2.8	ug/L	24	Standard
	Ce	140	216.7	24.0				ug/L	25	Standard
>	Tb	159	1024515.0	4.0				ug/L	939701	Standard
	Ho	165	13.3	57.3				ug/L	10	Standard
	Tl	203	467714.3	0.2	**50.8953**	1.428	2.8	ug/L	4	Standard
	Tl	205	419104.4	1.0	**52.5470**	1.541	2.9	ug/L	2	Standard
	Pb	206	302725.5	0.9	**50.7682**	1.716	3.4	ug/L	240	Standard
	Pb	207	257785.2	0.8	**48.6571**	1.672	3.4	ug/L	198	Standard
	Pb	208	1092150.5	0.9	**52.1642**	0.990	1.9	ug/L	883	Standard
	U	238	389261.3	0.3	**50.7252**	1.507	3.0	ug/L	3	Standard
>	Bi	209	584427.6	2.7				ug/L	523984	Standard

Sample ID: LCSW 43 WG573291-04
Report Date/Time: Wednesday, June 29, 2016 20:22:54
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	56.7	20.4	**-0.0027**	0.015	535.8	mg/L	37	Standard
K	39	35.0	42.9	**-0.0491**	0.026	52.4	mg/L	38	Standard
Ca	43	18.3	15.7	**-1.5405**	0.154	10.0	mg/L	10	Standard
Fe	54	173.6	39.1	**-0.0145**	0.044	301.9	mg/L	179	Standard
Fe	57	231.7	7.6	**-0.4064**	0.037	9.2	mg/L	335	Standard
Sc-1	45	71925.5	3.7				mg/L	61029	Standard
Cl	35	25598.9	2.5				ug/L	20195	Standard
Kr	83	1.0	100.0				ug/L	4	Standard
Br	81	1866.8	9.1				ug/L	1547	Standard
P	31	19263.9	1.4				ug/L	15286	Standard
S	34	2041.8	5.6				ug/L	1572	Standard
Sr	88	55.0	27.3				ug/L	73	Standard
C	12	76.7	37.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	6.7	173.2				mg/L	3	Standard
Dy	164	19.2	93.9				mg/L	6	Standard
Ho-1	165	13.3	57.3				mg/L	10	Standard
Er	166	16.7	124.9				mg/L	13	Standard
I	127	1405.1	22.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		106.142	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		117.021	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 43 WG573291-04
Report Date/Time: Wednesday, June 29, 2016 20:22:54
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
In	115		110.602
Sn	118		
Sb	123		
Ba	135		
Ce	140		
Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
Bi	209		111.535
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 43 WG573291-04
Report Date/Time: Wednesday, June 29, 2016 20:22:54
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401 WG573291-01
Sample Date/Time: Wednesday, June 29, 2016 20:23:48
Number of Replicates: 3
Autosampler Position: 412
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	81548.9	1.6				ug/L	74699	Standard
	Be	9	36.7	15.7	**0.0457**	0.006	12.1	ug/L	8	Standard
	Al	27	11596363.1	1.3	**72.2698**	0.564	0.8	ug/L	1493	Standard
	Sc	45	73884.7	1.3				ug/L	61029	Standard
	Ti	47	344.7	7.0	**0.7127**	0.059	8.2	ug/L	35	Standard
	V	51	2330.4	3.5	**0.0834**	0.015	17.8	ug/L	1435	Standard
	Cr	52	13306.8	2.8	**0.5995**	0.076	12.7	ug/L	7635	Standard
	Cr	53	1323.4	7.3	**0.9647**	0.110	11.4	ug/L	363	Standard
	Mn	55	1766896.9	0.7	**229.1909**	3.866	1.7	ug/L	747	Standard
	Co	59	11800.3	0.6	**1.5775**	0.021	1.4	ug/L	234	Standard
	Ni	60	3337.7	1.4	**1.4870**	0.026	1.7	ug/L	215	Standard
	Cu	65	729.0	2.5	**0.2502**	0.002	1.0	ug/L	104	Standard
	Zn	66	5843.1	1.1	**4.3791**	0.046	1.0	ug/L	363	Standard
>	**Ge**	72	548022.2	2.0				ug/L	495142	Standard
	As	75	368.2	8.1	**0.3569**	0.029	8.2	ug/L	-50	Standard
	Se	82	62.0	10.3	**0.3726**	0.060	16.1	ug/L	14	Standard
	Se-1	77	77.3	2.7	**0.3344**	0.023	6.7	ug/L	42	Standard
>	**Ga**	71	126.7	17.8				mg/L	42	Standard
	Rb	85	31524.4	1.4				ug/L	17	Standard
	Y	89	492847.4	3.2				ug/L	417098	Standard
>	**Rh**	103	60.0	14.4				ug/L	13	Standard
	Mo	98	169.7	8.5	**0.0488**	0.006	11.4	ug/L	11	Standard
	Ag	107	132.7	27.4	**0.0046**	0.005	112.7	ug/L	68	Standard
	Cd	111	39.0	30.5	**0.0081**	0.005	64.3	mg/L	2	Standard
	Cd	114	83.8	14.7	**0.0119**	0.002	19.5	ug/L	12	Standard
>	**In**	115	686408.4	2.0				ug/L	587441	Standard
	Sn	118	1350.1	6.2	**0.1033**	0.015	14.9	ug/L	497	Standard
	Sb	123	1853.5	25.3	**0.3348**	0.082	24.4	ug/L	78	Standard
	Ba	135	131659.9	0.9	**55.0560**	1.555	2.8	ug/L	24	Standard
	Ce	140	2260.2	0.8				ug/L	25	Standard
>	**Tb**	159	1046624.7	1.4				ug/L	939701	Standard
	Ho	165	75.0	29.1				ug/L	10	Standard
	Tl	203	216.3	9.4	**0.0226**	0.002	8.3	ug/L	4	Standard
	Tl	205	203.3	8.6	**0.0265**	0.002	7.2	ug/L	2	Standard
	Pb	206	535.3	4.4	**0.0217**	0.002	11.2	ug/L	240	Standard
	Pb	207	421.7	4.6	**0.0124**	0.003	25.0	ug/L	198	Standard
	Pb	208	1891.4	2.2	**0.0194**	0.001	4.1	ug/L	883	Standard
	U	238	92.7	15.8	**0.0080**	0.002	21.9	ug/L	3	Standard
>	**Bi**	209	565094.9	1.8				ug/L	523984	Standard

Sample ID: L1606097401 WG573291-01
Report Date/Time: Wednesday, June 29, 2016 20:26:05
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.0814**	27.845	173.2	mg/L	0	Standard
Mg	24	5874.5	4.9	**6.2016**	0.379	6.1	mg/L	37	Standard
K	39	696.7	5.3	**1.1574**	0.052	4.5	mg/L	38	Standard
Ca	43	108.3	19.2	**3.1206**	1.021	32.7	mg/L	10	Standard
Fe	54	752.2	10.4	**0.3885**	0.054	13.9	mg/L	179	Standard
Fe	57	410.0	11.6	**0.0774**	0.139	179.8	mg/L	335	Standard
Sc-1	45	73884.7	1.3				mg/L	61029	Standard
Cl	35	27603.8	2.8				ug/L	20195	Standard
Kr	83	3.0	66.7				ug/L	4	Standard
Br	81	11711.2	8.2				ug/L	1547	Standard
P	31	22032.7	3.4				ug/L	15286	Standard
S	34	2200.2	1.4				ug/L	1572	Standard
Sr	88	218.3	6.6				ug/L	73	Standard
C	12	103.3	43.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	3	Standard
Dy	164	96.2	10.7				mg/L	6	Standard
Ho-1	165	75.0	29.1				mg/L	10	Standard
Er	166	80.0	33.1				mg/L	13	Standard
I	127	173019.2	10.3				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		109.170	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		110.680	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606097401 WG573291-01
Report Date/Time: Wednesday, June 29, 2016 20:26:05
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		116.847
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.846
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606097401 WG573291-01
Report Date/Time: Wednesday, June 29, 2016 20:26:05
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097409S WG573291-05
Sample Date/Time: Wednesday, June 29, 2016 20:27:00
Number of Replicates: 3
Autosampler Position: 413
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77620.1	3.1				ug/L	74699	Standard
	Be	9	53947.6	2.9	**58.9852**	0.884	1.5	ug/L	8	Standard
	Al	27	10156417.9	2.3	**66.5068**	0.592	0.9	ug/L	1493	Standard
	Sc	45	69663.2	3.3				ug/L	61029	Standard
	Ti	47	217.3	7.1	**0.4124**	0.032	7.7	ug/L	35	Standard
	V	51	458159.6	2.1	**55.0030**	1.779	3.2	ug/L	1435	Standard
	Cr	52	452840.8	0.8	**54.5717**	1.179	2.2	ug/L	7635	Standard
	Cr	53	56488.5	2.6	**55.2901**	2.221	4.0	ug/L	363	Standard
	Mn	55	2082250.4	1.8	**271.9842**	8.205	3.0	ug/L	747	Standard
	Co	59	416796.8	1.5	**57.3937**	1.632	2.8	ug/L	234	Standard
	Ni	60	112326.5	0.8	**54.4621**	1.176	2.2	ug/L	215	Standard
	Cu	65	110417.6	0.6	**53.8701**	0.759	1.4	ug/L	104	Standard
	Zn	66	70959.5	1.2	**58.6468**	1.068	1.8	ug/L	363	Standard
>	**Ge**	72	544304.5	1.4				ug/L	495142	Standard
	As	75	66583.4	2.1	**54.1492**	1.789	3.3	ug/L	-50	Standard
	Se	82	6605.0	1.3	**54.6200**	1.423	2.6	ug/L	14	Standard
	Se-1	77	4636.4	0.8	**57.1316**	0.988	1.7	ug/L	42	Standard
>	**Ga**	71	90.0	19.2				mg/L	42	Standard
	Rb	85	30654.3	2.7				ug/L	17	Standard
	Y	89	497057.5	2.9				ug/L	417098	Standard
>	**Rh**	103	38.3	52.7				ug/L	13	Standard
	Mo	98	163.0	7.4	**0.0487**	0.002	4.2	ug/L	11	Standard
	Ag	107	371099.4	2.0	**50.7567**	1.341	2.6	ug/L	68	Standard
	Cd	111	118377.9	2.8	**50.4636**	1.124	2.2	mg/L	2	Standard
	Cd	114	281427.9	1.4	**48.9324**	1.490	3.0	ug/L	12	Standard
>	**In**	115	659308.0	4.2				ug/L	587441	Standard
	Sn	118	1090.0	7.1	**0.0721**	0.019	25.7	ug/L	497	Standard
	Sb	123	259620.0	1.8	**49.5761**	1.212	2.4	ug/L	78	Standard
	Ba	135	232712.5	1.7	**101.3728**	3.182	3.1	ug/L	24	Standard
	Ce	140	941.7	8.0				ug/L	25	Standard
>	**Tb**	159	982549.7	3.5				ug/L	939701	Standard
	Ho	165	31.7	24.1				ug/L	10	Standard
	Tl	203	453507.3	1.5	**53.4353**	0.634	1.2	ug/L	4	Standard
	Tl	205	415041.8	2.5	**56.3402**	0.862	1.5	ug/L	2	Standard
	Pb	206	296774.5	1.1	**53.8916**	0.589	1.1	ug/L	240	Standard
	Pb	207	252251.9	1.7	**51.5506**	0.378	0.7	ug/L	198	Standard
	Pb	208	1071354.8	2.3	**55.4200**	1.065	1.9	ug/L	883	Standard
	U	238	381907.7	2.6	**53.8787**	0.717	1.3	ug/L	3	Standard
>	**Bi**	209	539522.1	2.0				ug/L	523984	Standard

Sample ID: L1606097409S WG573291-05
Report Date/Time: Wednesday, June 29, 2016 20:29:17
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**17.4920**	30.288	173.2	mg/L	0	Standard
Mg	24	5339.3	3.4	**5.9734**	0.106	1.8	mg/L	37	Standard
K	39	681.7	10.6	**1.2038**	0.100	8.3	mg/L	38	Standard
Ca	43	128.3	25.3	**4.6217**	2.038	44.1	mg/L	10	Standard
Fe	54	542.4	5.6	**0.2649**	0.028	10.4	mg/L	179	Standard
Fe	57	371.7	15.7	**0.0334**	0.176	526.0	mg/L	335	Standard
> **Sc-1**	45	69663.2	3.3				mg/L	61029	Standard
Cl	35	28661.1	2.0				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	10680.4	5.1				ug/L	1547	Standard
P	31	21703.9	3.0				ug/L	15286	Standard
S	34	2123.5	4.7				ug/L	1572	Standard
Sr	88	153.3	22.2				ug/L	73	Standard
C	12	123.3	28.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	57.6	48.4				mg/L	6	Standard
Ho-1	165	31.7	24.1				mg/L	10	Standard
Er	166	50.0	91.7				mg/L	13	Standard
I	127	154285.8	7.0				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.911	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.929	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097409S WG573291-05
Report Date/Time: Wednesday, June 29, 2016 20:29:17
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.234
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		102.965
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097411SD WG573291-06
Sample Date/Time: Wednesday, June 29, 2016 20:30:11
Number of Replicates: 3
Autosampler Position: 414
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77405.6	1.5				ug/L	74699	Standard
	Be	9	55655.4	2.7	**61.0164**	1.576	2.6	ug/L	8	Standard
	Al	27	12086433.4	1.6	**79.3694**	1.985	2.5	ug/L	1493	Standard
	Sc	45	69212.8	2.6				ug/L	61029	Standard
	Ti	47	325.0	3.5	**0.6412**	0.031	4.8	ug/L	35	Standard
	V	51	466207.8	0.7	**53.8111**	1.021	1.9	ug/L	1435	Standard
	Cr	52	460631.7	1.4	**53.3535**	1.026	1.9	ug/L	7635	Standard
	Cr	53	56936.8	2.6	**53.5625**	1.349	2.5	ug/L	363	Standard
	Mn	55	2273187.5	1.8	**285.5286**	8.405	2.9	ug/L	747	Standard
	Co	59	417957.8	1.5	**55.3411**	1.481	2.7	ug/L	234	Standard
	Ni	60	114539.0	0.3	**53.3968**	0.715	1.3	ug/L	215	Standard
	Cu	65	111050.1	0.5	**52.0971**	0.842	1.6	ug/L	104	Standard
	Zn	66	69817.2	1.2	**55.4674**	1.299	2.3	ug/L	363	Standard
>	**Ge**	72	566018.8	1.2				ug/L	495142	Standard
	As	75	67402.1	1.8	**52.7097**	1.486	2.8	ug/L	-50	Standard
	Se	82	6620.3	2.0	**52.6418**	1.659	3.2	ug/L	14	Standard
	Se-1	77	4557.0	3.8	**53.9777**	2.700	5.0	ug/L	42	Standard
>	**Ga**	71	128.3	43.8				mg/L	42	Standard
	Rb	85	32409.6	0.5				ug/L	17	Standard
	Y	89	470382.0	1.0				ug/L	417098	Standard
>	**Rh**	103	61.7	24.8				ug/L	13	Standard
	Mo	98	196.1	5.9	**0.0618**	0.004	5.9	ug/L	11	Standard
	Ag	107	374581.1	0.3	**52.7275**	0.291	0.6	ug/L	68	Standard
	Cd	111	118658.6	1.5	**52.0656**	0.397	0.8	mg/L	2	Standard
	Cd	114	286274.0	2.6	**51.2138**	0.959	1.9	ug/L	12	Standard
>	**In**	115	640194.5	0.7				ug/L	587441	Standard
	Sn	118	1251.7	9.2	**0.1018**	0.017	16.3	ug/L	497	Standard
	Sb	123	263336.7	0.4	**51.7531**	0.327	0.6	ug/L	78	Standard
	Ba	135	250997.3	0.8	**112.5129**	0.934	0.8	ug/L	24	Standard
	Ce	140	2000.1	6.5				ug/L	25	Standard
>	**Tb**	159	984120.1	1.6				ug/L	939701	Standard
	Ho	165	48.3	11.9				ug/L	10	Standard
	Tl	203	459854.1	0.1	**53.6772**	0.823	1.5	ug/L	4	Standard
	Tl	205	414544.5	0.5	**55.7557**	1.089	2.0	ug/L	2	Standard
	Pb	206	298100.5	1.1	**53.6266**	1.109	2.1	ug/L	240	Standard
	Pb	207	253941.0	0.3	**51.4145**	0.875	1.7	ug/L	198	Standard
	Pb	208	1061848.2	0.5	**54.4146**	0.782	1.4	ug/L	883	Standard
	U	238	385197.1	0.6	**53.8449**	1.070	2.0	ug/L	3	Standard
>	**Bi**	209	544631.3	1.4				ug/L	523984	Standard

Sample ID: L1606097411SD WG573291-06
Report Date/Time: Wednesday, June 29, 2016 20:32:28
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	5929.5	2.1	**6.6847**	0.075	1.1	mg/L	37	Standard
K	39	801.7	4.2	**1.4479**	0.027	1.8	mg/L	38	Standard
Ca	43	125.0	31.2	**4.4423**	2.177	49.0	mg/L	10	Standard
Fe	54	841.7	6.4	**0.4916**	0.050	10.2	mg/L	179	Standard
Fe	57	441.7	5.4	**0.2489**	0.040	15.9	mg/L	335	Standard
> **Sc-1**	45	69212.8	2.6				mg/L	61029	Standard
Cl	35	28869.5	0.9				ug/L	20195	Standard
Kr	83	3.7	56.8				ug/L	4	Standard
Br	81	12511.8	0.5				ug/L	1547	Standard
P	31	22511.8	6.4				ug/L	15286	Standard
S	34	2320.2	4.3				ug/L	1572	Standard
Sr	88	170.0	8.8				ug/L	73	Standard
C	12	80.0	25.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	3	Standard
Dy	164	93.8	22.7				mg/L	6	Standard
Ho-1	165	48.3	11.9				mg/L	10	Standard
Er	166	60.0	57.7				mg/L	13	Standard
I	127	186712.7	7.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.624	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.314	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097411SD WG573291-06
Report Date/Time: Wednesday, June 29, 2016 20:32:28
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.980
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.940
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606097411SD WG573291-06
Report Date/Time: Wednesday, June 29, 2016 20:32:28
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401PS WG573790-03
Sample Date/Time: Wednesday, June 29, 2016 20:33:23
Number of Replicates: 3
Autosampler Position: 415
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77767.4	1.4				ug/L	74699	Standard
	Be	9	54620.0	3.0	**59.5914**	1.002	1.7	ug/L	8	Standard
	Al	27	11452990.5	1.0	**74.8575**	1.594	2.1	ug/L	1493	Standard
	Sc	45	70666.1	1.7				ug/L	61029	Standard
	Ti	47	337.7	3.5	**0.6996**	0.031	4.4	ug/L	35	Standard
	V	51	448357.3	1.6	**53.6764**	0.056	0.1	ug/L	1435	Standard
	Cr	52	455460.2	1.4	**54.7521**	0.846	1.5	ug/L	7635	Standard
	Cr	53	57375.1	1.7	**56.0052**	0.200	0.4	ug/L	363	Standard
	Mn	55	2154709.7	0.5	**280.7294**	3.122	1.1	ug/L	747	Standard
	Co	59	404569.2	0.4	**55.5661**	0.653	1.2	ug/L	234	Standard
	Ni	60	112633.8	1.0	**54.4757**	0.921	1.7	ug/L	215	Standard
	Cu	65	109355.9	1.1	**53.2239**	1.011	1.9	ug/L	104	Standard
	Zn	66	70498.3	1.6	**58.1166**	0.789	1.4	ug/L	363	Standard
>	**Ge**	72	545627.2	1.6				ug/L	495142	Standard
	As	75	64850.8	0.9	**52.6029**	0.345	0.7	ug/L	-50	Standard
	Se	82	6485.5	1.8	**53.4847**	0.149	0.3	ug/L	14	Standard
	Se-1	77	4411.0	1.9	**54.1848**	0.787	1.5	ug/L	42	Standard
>	**Ga**	71	161.7	1.8				mg/L	42	Standard
	Rb	85	30524.1	1.7				ug/L	17	Standard
	Y	89	474875.4	0.7				ug/L	417098	Standard
>	**Rh**	103	60.0	16.7				ug/L	13	Standard
	Mo	98	163.5	8.5	**0.0491**	0.004	8.3	ug/L	11	Standard
	Ag	107	365805.1	1.5	**50.1795**	1.312	2.6	ug/L	68	Standard
	Cd	111	117534.8	0.9	**50.2525**	0.249	0.5	mg/L	2	Standard
	Cd	114	285376.1	1.5	**49.7475**	0.497	1.0	ug/L	12	Standard
>	**In**	115	657058.7	1.4				ug/L	587441	Standard
	Sn	118	1523.4	8.4	**0.1383**	0.017	12.1	ug/L	497	Standard
	Sb	123	259036.8	1.0	**49.6101**	1.139	2.3	ug/L	78	Standard
	Ba	135	239792.4	1.1	**104.7386**	1.629	1.6	ug/L	24	Standard
	Ce	140	2208.5	2.2				ug/L	25	Standard
>	**Tb**	159	1019758.9	1.4				ug/L	939701	Standard
	Ho	165	71.7	21.3				ug/L	10	Standard
	Tl	203	456925.9	1.0	**51.8248**	0.434	0.8	ug/L	4	Standard
	Tl	205	410087.8	1.3	**53.5969**	1.082	2.0	ug/L	2	Standard
	Pb	206	289464.5	1.3	**50.5998**	1.170	2.3	ug/L	240	Standard
	Pb	207	248071.4	0.9	**48.8073**	1.130	2.3	ug/L	198	Standard
	Pb	208	1050860.8	0.8	**52.3325**	1.144	2.2	ug/L	883	Standard
	U	238	380364.0	0.7	**51.6678**	1.019	2.0	ug/L	3	Standard
>	**Bi**	209	560465.1	1.6				ug/L	523984	Standard

Sample ID: L1606097401PS WG573790-03
Report Date/Time: Wednesday, June 29, 2016 20:35:40
Page 1

Approved: July 04, 2016

Na	23	1.7	173.2	**16.2901**	28.207	173.2	mg/L	0	Standard
Mg	24	5812.8	3.2	**6.4153**	0.191	3.0	mg/L	37	Standard
K	39	706.7	5.8	**1.2347**	0.072	5.8	mg/L	38	Standard
Ca	43	108.3	32.4	**3.3848**	1.922	56.8	mg/L	10	Standard
Fe	54	723.0	16.6	**0.3920**	0.096	24.5	mg/L	179	Standard
Fe	57	435.0	5.7	**0.2027**	0.073	35.8	mg/L	335	Standard
> **Sc-1**	45	70666.1	1.7				mg/L	61029	Standard
Cl	35	28625.7	0.6				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	11964.7	4.0				ug/L	1547	Standard
P	31	22167.9	1.3				ug/L	15286	Standard
S	34	2268.5	2.8				ug/L	1572	Standard
Sr	88	201.7	11.5				ug/L	73	Standard
C	12	93.3	34.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	3	Standard
Dy	164	120.1	22.8				mg/L	6	Standard
Ho-1	165	71.7	21.3				mg/L	10	Standard
Er	166	66.7	60.6				mg/L	13	Standard
I	127	179600.9	7.9				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		104.108	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.196	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097401PS WG573790-03
Report Date/Time: Wednesday, June 29, 2016 20:35:40
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		111.851
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.962
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606097401PS WG573790-03
Report Date/Time: Wednesday, June 29, 2016 20:35:40
Page 3

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401SDL WG573790-04
Sample Date/Time: Wednesday, June 29, 2016 20:36:34
Number of Replicates: 3
Autosampler Position: 416
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	77070.7	3.6				ug/L	74699	Standard
	Be	9	45.0	33.3	**0.0568**	0.015	26.1	ug/L	8	Standard
	Al	27	2260304.8	1.2	**14.9037**	0.448	3.0	ug/L	1493	Standard
	Sc	45	68278.6	3.1				ug/L	61029	Standard
	Ti	47	83.0	9.9	**0.0892**	0.019	21.8	ug/L	35	Standard
	V	51	2192.5	3.4	**0.0673**	0.008	12.3	ug/L	1435	Standard
	Cr	52	10144.1	1.9	**0.2156**	0.020	9.5	ug/L	7635	Standard
	Cr	53	686.7	4.8	**0.3429**	0.034	10.1	ug/L	363	Standard
	Mn	55	355711.8	1.3	**46.1130**	0.456	1.0	ug/L	747	Standard
	Co	59	2492.2	1.8	**0.3045**	0.007	2.2	ug/L	234	Standard
	Ni	60	828.7	2.9	**0.2769**	0.011	3.8	ug/L	215	Standard
	Cu	65	280.0	0.7	**0.0325**	0.001	2.6	ug/L	104	Standard
	Zn	66	2517.5	2.7	**1.6325**	0.059	3.6	ug/L	363	Standard
>	**Ge**	72	546825.8	0.3				ug/L	495142	Standard
	As	75	65.5	45.7	**0.1126**	0.024	21.6	ug/L	-50	Standard
	Se	82	28.6	15.9	**0.0977**	0.038	38.5	ug/L	14	Standard
	Se-1	77	61.3	12.0	**0.1380**	0.090	65.2	ug/L	42	Standard
>	**Ga**	71	98.3	5.9				mg/L	42	Standard
	Rb	85	6301.3	2.8				ug/L	17	Standard
	Y	89	489576.3	2.2				ug/L	417098	Standard
>	**Rh**	103	21.7	48.0				ug/L	13	Standard
	Mo	98	47.0	21.8	**0.0100**	0.004	35.1	ug/L	11	Standard
	Ag	107	179.0	15.1	**0.0115**	0.004	30.5	ug/L	68	Standard
	Cd	111	35.6	11.4	**0.0072**	0.002	22.2	mg/L	2	Standard
	Cd	114	102.0	28.8	**0.0155**	0.005	31.8	ug/L	12	Standard
>	**In**	115	661161.0	0.8				ug/L	587441	Standard
	Sn	118	795.0	15.7	**0.0261**	0.018	68.8	ug/L	497	Standard
	Sb	123	2322.3	23.7	**0.4382**	0.107	24.4	ug/L	78	Standard
	Ba	135	26230.7	1.3	**11.3705**	0.051	0.4	ug/L	24	Standard
	Ce	140	480.0	13.3				ug/L	25	Standard
>	**Tb**	159	1000122.3	1.4				ug/L	939701	Standard
	Ho	165	26.7	10.8				ug/L	10	Standard
	Tl	203	156.0	22.9	**0.0163**	0.004	24.4	ug/L	4	Standard
	Tl	205	150.0	15.3	**0.0202**	0.003	14.3	ug/L	2	Standard
	Pb	206	403.0	8.6	**0.0007**	0.006	871.4	ug/L	240	Standard
	Pb	207	342.3	5.9	**-0.0013**	0.004	288.2	ug/L	198	Standard
	Pb	208	1497.0	5.1	**0.0020**	0.004	181.5	ug/L	883	Standard
	U	238	114.0	13.2	**0.0113**	0.002	17.4	ug/L	3	Standard
>	**Bi**	209	549941.1	1.0				ug/L	523984	Standard

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:38:51
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	1213.4	3.4	**1.3365**	0.084	6.3	mg/L	37	Standard
K	39	180.0	11.1	**0.2425**	0.050	20.5	mg/L	38	Standard
Ca	43	46.7	22.3	**0.1064**	0.547	514.2	mg/L	10	Standard
Fe	54	286.0	10.0	**0.0782**	0.017	21.6	mg/L	179	Standard
Fe	57	306.7	10.9	**-0.1437**	0.079	55.0	mg/L	335	Standard
> **Sc-1**	45	68278.6	3.1				mg/L	61029	Standard
Cl	35	28506.1	0.3				ug/L	20195	Standard
Kr	83	1.7	34.6				ug/L	4	Standard
Br	81	3947.2	6.9				ug/L	1547	Standard
P	31	19479.2	3.3				ug/L	15286	Standard
S	34	2088.5	5.3				ug/L	1572	Standard
Sr	88	101.7	24.8				ug/L	73	Standard
C	12	60.0	60.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	3	Standard
Dy	164	52.5	9.9				mg/L	6	Standard
Ho-1	165	26.7	10.8				mg/L	10	Standard
Er	166	16.7	69.3				mg/L	13	Standard
I	127	39316.0	2.7				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.175	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.438	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:38:51
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		112.549
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.954
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: L1606097401SDL WG573790-04
Sample Date/Time: Wednesday, June 29, 2016 20:39:46
Number of Replicates: 3
Autosampler Position: 417
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	76547.9	2.1				ug/L	74699	Standard
	Be	9	28.3	27.0	**0.0389**	0.008	20.5	ug/L	8	Standard
	Al	27	465099.1	0.7	**3.0756**	0.044	1.4	ug/L	1493	Standard
	Sc	45	68169.7	0.9				ug/L	61029	Standard
	Ti	47	36.7	6.9	**-0.0243**	0.006	24.2	ug/L	35	Standard
	V	51	2063.3	3.2	**0.0438**	0.007	16.8	ug/L	1435	Standard
	Cr	52	9662.1	0.8	**0.1188**	0.007	6.3	ug/L	7635	Standard
	Cr	53	523.3	10.2	**0.1664**	0.051	30.7	ug/L	363	Standard
	Mn	55	72938.2	2.1	**9.0325**	0.175	1.9	ug/L	747	Standard
	Co	59	731.7	15.3	**0.0598**	0.015	24.7	ug/L	234	Standard
	Ni	60	309.0	4.2	**0.0207**	0.006	30.5	ug/L	215	Standard
	Cu	65	310.7	4.3	**0.0426**	0.006	14.7	ug/L	104	Standard
	Zn	66	2601.2	2.1	**1.6324**	0.044	2.7	ug/L	363	Standard
>	**Ge**	72	565002.4	0.2				ug/L	495142	Standard
	As	75	-33.8	92.8	**0.0331**	0.025	74.3	ug/L	-50	Standard
	Se	82	20.9	22.0	**0.0285**	0.036	128.0	ug/L	14	Standard
	Se-1	77	57.7	17.5	**0.0697**	0.122	174.5	ug/L	42	Standard
>	**Ga**	71	58.3	9.9				mg/L	42	Standard
	Rb	85	1335.1	7.2				ug/L	17	Standard
	Y	89	490772.8	1.5				ug/L	417098	Standard
>	**Rh**	103	15.0	57.7				ug/L	13	Standard
	Mo	98	24.2	10.4	**0.0025**	0.001	32.4	ug/L	11	Standard
	Ag	107	143.3	39.3	**0.0071**	0.008	112.3	ug/L	68	Standard
	Cd	111	11.6	65.7	**-0.0029**	0.003	116.5	mg/L	2	Standard
	Cd	114	25.4	61.7	**0.0024**	0.003	115.9	ug/L	12	Standard
>	**In**	115	647709.2	0.5				ug/L	587441	Standard
	Sn	118	543.3	5.3	**-0.0103**	0.004	42.8	ug/L	497	Standard
	Sb	123	784.5	24.6	**0.1481**	0.037	24.8	ug/L	78	Standard
	Ba	135	5327.9	0.7	**2.3449**	0.015	0.6	ug/L	24	Standard
	Ce	140	110.0	7.9				ug/L	25	Standard
>	**Tb**	159	986009.2	0.5				ug/L	939701	Standard
	Ho	165	10.0	100.0				ug/L	10	Standard
	Tl	203	50.7	34.9	**0.0041**	0.002	51.6	ug/L	4	Standard
	Tl	205	46.7	71.3	**0.0064**	0.005	70.3	ug/L	2	Standard
	Pb	206	376.3	19.7	**-0.0043**	0.014	323.1	ug/L	240	Standard
	Pb	207	290.7	20.1	**-0.0119**	0.012	103.4	ug/L	198	Standard
	Pb	208	1284.0	11.0	**-0.0091**	0.008	86.4	ug/L	883	Standard
	U	238	44.0	79.9	**0.0017**	0.005	294.8	ug/L	3	Standard
>	**Bi**	209	552760.9	1.2				ug/L	523984	Standard

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:42:03
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	246.7	9.1	**0.2201**	0.028	12.6	mg/L	37	Standard
K	39	50.0	40.0	**-0.0154**	0.039	250.4	mg/L	38	Standard
Ca	43	20.0	43.3	**-1.3952**	0.477	34.2	mg/L	10	Standard
Fe	54	182.5	25.5	**0.0004**	0.036	9653.1	mg/L	179	Standard
Fe	57	256.7	8.8	**-0.2933**	0.068	23.1	mg/L	335	Standard
Sc-1	45	68169.7	0.9				mg/L	61029	Standard
Cl	35	28987.0	1.4				ug/L	20195	Standard
Kr	83	1.7	69.3				ug/L	4	Standard
Br	81	2360.2	9.1				ug/L	1547	Standard
P	31	19090.4	2.0				ug/L	15286	Standard
S	34	2190.2	0.6				ug/L	1572	Standard
Sr	88	73.3	27.6				ug/L	73	Standard
C	12	46.7	44.6				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	22.7	102.4				mg/L	6	Standard
Ho-1	165	10.0	100.0				mg/L	10	Standard
Er	166	13.3	114.6				mg/L	13	Standard
I	127	10453.6	7.5				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		102.475	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		114.109	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606097401SDL WG573790-04
Report Date/Time: Wednesday, June 29, 2016 20:42:03
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		110.259
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.492
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Wednesday, June 29, 2016 20:42:59
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	73968.6	3.7				ug/L	74699	Standard
	Be	9	51786.8	4.7	**59.3953**	0.618	1.0	ug/L	8	Standard
	Al	27	8138418.2	2.1	**55.9353**	1.016	1.8	ug/L	1493	Standard
	Sc	45	66422.0	2.7				ug/L	61029	Standard
	Ti	47	44367.5	1.4	**105.9251**	2.261	2.1	ug/L	35	Standard
	V	51	448161.7	0.9	**53.5504**	0.563	1.1	ug/L	1435	Standard
	Cr	52	438098.2	0.5	**52.5206**	0.674	1.3	ug/L	7635	Standard
	Cr	53	53648.2	1.8	**52.2574**	1.658	3.2	ug/L	363	Standard
	Mn	55	415062.2	0.8	**53.8560**	1.244	2.3	ug/L	747	Standard
	Co	59	407202.7	0.3	**55.8201**	1.033	1.9	ug/L	234	Standard
	Ni	60	107214.7	0.8	**51.7442**	0.748	1.4	ug/L	215	Standard
	Cu	65	106835.8	1.9	**51.8884**	1.090	2.1	ug/L	104	Standard
	Zn	66	63121.9	0.7	**51.8882**	0.831	1.6	ug/L	363	Standard
>	Ge	72	546729.2	1.7				ug/L	495142	Standard
	As	75	63950.8	1.3	**51.7792**	1.311	2.5	ug/L	-50	Standard
	Se	82	6522.2	2.4	**53.7001**	1.978	3.7	ug/L	14	Standard
	Se-1	77	4462.7	1.0	**54.7224**	0.944	1.7	ug/L	42	Standard
>	Ga	71	48.3	11.9				mg/L	42	Standard
	Rb	85	921.7	10.0				ug/L	17	Standard
	Y	89	469877.0	1.2				ug/L	417098	Standard
>	Rh	103	31.7	45.6				ug/L	13	Standard
	Mo	98	306064.1	0.6	**105.7488**	1.650	1.6	ug/L	11	Standard
	Ag	107	358011.4	0.7	**50.5757**	0.461	0.9	ug/L	68	Standard
	Cd	111	116953.2	1.0	**51.5035**	0.389	0.8	mg/L	2	Standard
	Cd	114	288128.0	0.1	**51.7395**	0.785	1.5	ug/L	12	Standard
>	In	115	637946.0	1.6				ug/L	587441	Standard
	Sn	118	327214.8	2.0	**51.4831**	0.238	0.5	ug/L	497	Standard
	Sb	123	254525.5	0.5	**50.2034**	0.722	1.4	ug/L	78	Standard
	Ba	135	108828.5	1.0	**48.9489**	0.332	0.7	ug/L	24	Standard
	Ce	140	90.0	29.4				ug/L	25	Standard
>	Tb	159	958314.6	1.8				ug/L	939701	Standard
	Ho	165	11.7	49.5				ug/L	10	Standard
	Tl	203	447416.0	0.6	**53.4811**	1.019	1.9	ug/L	4	Standard
	Tl	205	409322.9	1.6	**56.3810**	1.712	3.0	ug/L	2	Standard
	Pb	206	283417.6	0.7	**52.2045**	0.689	1.3	ug/L	240	Standard
	Pb	207	252263.1	0.7	**52.2984**	0.513	1.0	ug/L	198	Standard
	Pb	208	1062273.6	1.1	**55.7538**	1.546	2.8	ug/L	883	Standard
	U	238	385444.7	0.9	**55.1725**	1.082	2.0	ug/L	3	Standard
>	Bi	209	531871.1	1.6				ug/L	523984	Standard

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	4377.3	2.2	**5.1285**	0.149	2.9	mg/L	37	Standard
K	39	2180.2	1.8	**4.3148**	0.109	2.5	mg/L	38	Standard
Ca	43	118.3	10.6	**4.3366**	0.588	13.6	mg/L	10	Standard
Fe	54	6990.5	4.3	**5.3088**	0.233	4.4	mg/L	179	Standard
Fe	57	2083.5	2.8	**5.4386**	0.352	6.5	mg/L	335	Standard
> **Sc-1**	45	66422.0	2.7				mg/L	61029	Standard
Cl	35	29309.7	0.2				ug/L	20195	Standard
Kr	83	2.3	24.7				ug/L	4	Standard
Br	81	2156.8	3.9				ug/L	1547	Standard
P	31	21223.2	2.6				ug/L	15286	Standard
S	34	2665.2	2.0				ug/L	1572	Standard
Sr	88	71.7	8.1				ug/L	73	Standard
C	12	56.7	40.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	19.8	1.4				mg/L	6	Standard
Ho-1	165	11.7	49.5				mg/L	10	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	2623.6	23.6				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	118.791		
Al	27	111.871		
Sc	45			
Ti	47	105.925		
V	51	107.101		
Cr	52	105.041		
Cr	53			
Mn	55	107.712		
Co	59	111.640		
Ni	60	103.488		
Cu	65	103.777		
Zn	66	103.776		
> Ge	72		110.419	
As	75	103.558		
Se	82	107.400		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 2

Approved: July 04, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	105.749		
Ag	107	101.151		
Cd	111	103.007		
Cd	114			
In	115		108.597	
Sn	118	102.966		
Sb	123	100.407		
Ba	135	97.898		
Ce	140			
Tb	159			
Ho	165			
Tl	203	106.962		
Tl	205			
Pb	206			
Pb	207			
Pb	208	111.508		
U	238	110.345		
Bi	209		101.505	
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 6	Be	9	
QC Std 6	Al	27	
QC Std 6	Co	59	

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 3

Approved: July 04, 2016

QC Std 6	Pb	208
QC Std 6	U	238

Sample ID: QC Std 6
Report Date/Time: Wednesday, June 29, 2016 20:45:16
Page 4

Approved: July 04, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Wednesday, June 29, 2016 20:46:11
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: BKT Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	75383.8	1.3				ug/L	74699	Standard
	Be	9	28.3	20.4	**0.0395**	0.006	15.9	ug/L	8	Standard
	Al	27	2638.6	20.0	**0.0044**	0.003	78.6	ug/L	1493	Standard
	Sc	45	68713.9	3.5				ug/L	61029	Standard
	Ti	47	37.7	13.4	**-0.0232**	0.011	49.4	ug/L	35	Standard
	V	51	1533.2	3.6	**-0.0202**	0.008	41.3	ug/L	1435	Standard
	Cr	52	7879.4	2.2	**-0.1054**	0.029	27.9	ug/L	7635	Standard
	Cr	53	448.3	16.3	**0.0885**	0.063	71.1	ug/L	363	Standard
	Mn	55	901.4	5.8	**-0.0363**	0.007	18.9	ug/L	747	Standard
	Co	59	283.3	10.0	**-0.0002**	0.004	1542.3	ug/L	234	Standard
	Ni	60	186.3	2.6	**-0.0379**	0.003	8.3	ug/L	215	Standard
	Cu	65	132.7	3.4	**-0.0421**	0.001	3.4	ug/L	104	Standard
	Zn	66	415.3	5.4	**-0.1263**	0.014	10.9	ug/L	363	Standard
>	**Ge**	72	573402.3	1.2				ug/L	495142	Standard
	As	75	-34.2	13.3	**0.0332**	0.003	10.5	ug/L	-50	Standard
	Se	82	17.2	23.7	**-0.0025**	0.031	1279.2	ug/L	14	Standard
	Se-1	77	54.3	12.3	**0.0204**	0.082	402.5	ug/L	42	Standard
>	**Ga**	71	35.0	37.8				mg/L	42	Standard
	Rb	85	30.0	28.9				ug/L	17	Standard
	Y	89	465691.4	1.8				ug/L	417098	Standard
>	**Rh**	103	6.7	43.3				ug/L	13	Standard
	Mo	98	268.8	25.2	**0.0874**	0.025	29.0	ug/L	11	Standard
	Ag	107	132.7	5.0	**0.0058**	0.001	22.7	ug/L	68	Standard
	Cd	111	15.2	35.4	**-0.0012**	0.003	209.6	mg/L	2	Standard
	Cd	114	58.4	20.0	**0.0084**	0.002	23.7	ug/L	12	Standard
>	**In**	115	638175.3	2.2				ug/L	587441	Standard
	Sn	118	866.7	15.8	**0.0416**	0.019	45.6	ug/L	497	Standard
	Sb	123	616.5	30.1	**0.1169**	0.034	29.1	ug/L	78	Standard
	Ba	135	50.7	18.3	**0.0068**	0.005	67.0	ug/L	24	Standard
	Ce	140	28.3	27.0				ug/L	25	Standard
>	**Tb**	159	990005.2	3.0				ug/L	939701	Standard
	Ho	165	6.7	86.6				ug/L	10	Standard
	Tl	203	92.0	24.5	**0.0086**	0.003	32.5	ug/L	4	Standard
	Tl	205	76.7	43.4	**0.0102**	0.005	44.4	ug/L	2	Standard
	Pb	206	305.7	10.6	**-0.0179**	0.007	38.8	ug/L	240	Standard
	Pb	207	274.7	5.9	**-0.0162**	0.004	23.3	ug/L	198	Standard
	Pb	208	1058.7	7.3	**-0.0216**	0.005	23.5	ug/L	883	Standard
	U	238	60.3	25.7	**0.0037**	0.002	61.2	ug/L	3	Standard
>	**Bi**	209	564302.2	2.4				ug/L	523984	Standard

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:48:28
Page 1

Approved: July 04, 2016

Na	23	0.0		**0.0050**	0.000	0.0	mg/L	0	Standard
Mg	24	35.0	37.8	**-0.0253**	0.014	54.1	mg/L	37	Standard
K	39	45.0	19.2	**-0.0261**	0.014	54.4	mg/L	38	Standard
Ca	43	26.7	39.0	**-1.0220**	0.608	59.5	mg/L	10	Standard
Fe	54	162.5	17.8	**-0.0163**	0.018	110.8	mg/L	179	Standard
Fe	57	258.3	8.1	**-0.2948**	0.047	15.9	mg/L	335	Standard
> **Sc-1**	45	68713.9	3.5				mg/L	61029	Standard
Cl	35	28935.6	0.2				ug/L	20195	Standard
Kr	83	2.0	50.0				ug/L	4	Standard
Br	81	1740.1	12.2				ug/L	1547	Standard
P	31	21446.8	2.8				ug/L	15286	Standard
S	34	2463.5	5.1				ug/L	1572	Standard
Sr	88	71.7	17.6				ug/L	73	Standard
C	12	43.3	13.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	3	Standard
Dy	164	12.9	115.2				mg/L	6	Standard
Ho-1	165	6.7	86.6				mg/L	10	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	2023.5	28.1				mg/L	1392	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.806	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:48:28
Page 2

Approved: July 04, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.636
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.694
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Wednesday, June 29, 2016 20:48:28
Page 3

Approved: July 04, 2016

MassCal File Name

Mass Calibration File Name Default.tun
MassCal File Path C:\NexIONData\MassCal\Default.tun
Peak Search Window: 1.00

Sample Information

Sample Date/Time: Tuesday, July 05, 2016 09:04:54

Mass Calibration and Resolution

Analyte	E Mass	Meas Mass	Mass C DAC Val	Res DAC Value	Meas Peak W	Custom Res
Li	7.016	6.975	1321	2024	0.708	
Mg	23.985	23.975	4501	2020	0.692	
Co	58.933	58.925	11690	2022	0.690	
In	114.904	114.925	22863	2028	0.687	
U	238.050	238.075	47461	2043	0.681	

Relative Std. Dev.

Mass	Meas. Intens. RSD
5.525	4.916
5.575	6.266
5.625	2.810
5.675	3.120
5.725	3.210
5.775	1.700
5.825	3.914
5.875	4.024
5.925	2.609
5.975	1.860
6.025	2.733
6.075	1.681
6.125	3.202
6.175	4.353
6.225	8.355
6.275	67.748
6.325	136.931
6.375	108.327
6.425	81.650
6.475	34.993
6.525	7.996
6.575	6.006
6.625	4.810
6.675	2.293
6.725	1.911
6.775	1.582
6.825	2.290

Report Date/Time: Tuesday, July 05, 2016 09:08:05
Page 1

Approved: July 06, 2016

6.875	1.932
6.925	2.048
6.975	1.464
7.025	2.377
7.075	1.360
7.125	1.136
7.175	1.498
7.225	2.429
7.275	4.978
7.325	12.811
7.375	50.000
7.425	43.853
7.475	59.266
7.525	60.858
7.575	71.261
7.625	59.266
7.675	37.268
7.725	71.261
7.775	61.237
7.825	46.481
7.875	119.523
7.925	70.711
7.975	63.888
8.025	141.421
8.075	70.711
8.125	104.583
8.175	59.266
8.225	37.268
8.275	39.123
8.325	20.328
8.375	70.711
8.425	72.436
8.475	37.268
22.525	223.607
22.575	67.420
22.625	59.894
22.675	54.149
22.725	52.322
22.775	28.395
22.825	9.448
22.875	13.869
22.925	29.601
22.975	15.049
23.025	24.549
23.075	52.605
23.125	27.664
23.175	25.514

Report Date/Time: Tuesday, July 05, 2016 09:08:05
Page 2

Approved: July 06, 2016

23.225	34.035
23.275	38.083
23.325	53.452
23.375	19.218
23.425	19.357
23.475	24.628
23.525	2.185
23.575	1.481
23.625	1.660
23.675	1.271
23.725	1.432
23.775	1.289
23.825	0.651
23.875	1.352
23.925	1.696
23.975	2.349
24.025	2.044
24.075	1.668
24.125	1.409
24.175	1.436
24.225	1.673
24.275	2.430
24.325	13.116
24.375	42.127
24.425	47.222
24.475	3.626
24.525	3.161
24.575	4.479
24.625	0.720
24.675	2.084
24.725	1.446
24.775	0.782
24.825	1.059
24.875	0.986
24.925	1.507
24.975	1.323
25.025	0.995
25.075	1.253
25.125	1.235
25.175	0.877
25.225	3.966
25.275	6.789
25.325	68.148
25.375	62.630
25.425	18.054
25.475	22.592
57.525	8.108

Report Date/Time: Tuesday, July 05, 2016 09:08:05
Page 3

Approved: July 06, 2016

57.575	5.819
57.625	3.555
57.675	1.890
57.725	0.683
57.775	1.819
57.825	0.687
57.875	2.247
57.925	1.417
57.975	0.754
58.025	1.690
58.075	1.590
58.125	2.082
58.175	1.933
58.225	2.633
58.275	11.735
58.325	28.571
58.375	45.644
58.425	13.451
58.475	8.093
58.525	7.046
58.575	2.359
58.625	1.975
58.675	1.980
58.725	2.667
58.775	1.582
58.825	1.860
58.875	1.841
58.925	1.155
58.975	1.132
59.025	2.557
59.075	1.177
59.125	2.415
59.175	1.508
59.225	2.089
59.275	23.436
59.325	97.183
59.375	43.241
59.425	44.628
59.475	17.563
59.525	9.370
59.575	3.904
59.625	4.547
59.675	4.189
59.725	1.832
59.775	1.847
59.825	3.828
59.875	2.178

Report Date/Time: Tuesday, July 05, 2016 09:08:05
Page 4

Approved: July 06, 2016

59.925	2.815
59.975	2.630
60.025	3.394
60.075	2.243
60.125	2.466
60.175	2.169
60.225	6.650
60.275	28.144
60.325	47.507
60.375	26.307
60.425	117.656
60.475	54.294
113.525	7.903
113.575	5.232
113.625	3.878
113.675	1.766
113.725	3.345
113.775	1.670
113.825	2.803
113.875	1.444
113.925	2.208
113.975	1.811
114.025	2.067
114.075	3.273
114.125	1.543
114.175	2.410
114.225	3.081
114.275	5.715
114.325	11.478
114.375	31.623
114.425	20.012
114.475	14.173
114.525	2.838
114.575	2.225
114.625	2.421
114.675	2.092
114.725	1.273
114.775	1.395
114.825	1.091
114.875	1.993
114.925	1.636
114.975	1.467
115.025	2.093
115.075	0.811
115.125	1.853
115.175	2.044
115.225	2.848

Report Date/Time: Tuesday, July 05, 2016 09:08:05
Page 5

Approved: July 06, 2016

115.275	5.340
115.325	14.538
115.375	66.144
115.425	53.926
115.475	49.311
115.525	13.098
115.575	6.772
115.625	6.176
115.675	6.649
115.725	2.929
115.775	2.384
115.825	1.377
115.875	2.818
115.925	2.791
115.975	4.600
116.025	1.704
116.075	3.015
116.125	5.054
116.175	3.417
116.225	10.059
116.275	14.933
116.325	39.316
116.375	80.889
116.425	63.738
116.475	62.828
236.525	223.607
236.575	24.686
236.625	26.702
236.675	26.438
236.725	35.460
236.775	14.582
236.825	54.128
236.875	30.292
236.925	43.971
236.975	45.138
237.025	21.123
237.075	22.356
237.125	16.110
237.175	24.491
237.225	31.558
237.275	15.972
237.325	39.891
237.375	21.066
237.425	11.877
237.475	32.439
237.525	35.561
237.575	11.436

Report Date/Time: Tuesday, July 05, 2016 09:08:05
Page 6

Approved: July 06, 2016

237.625	7.182
237.675	3.176
237.725	2.457
237.775	0.976
237.825	2.661
237.875	0.713
237.925	0.911
237.975	1.333
238.025	0.816
238.075	1.224
238.125	0.903
238.175	1.478
238.225	1.941
238.275	0.792
238.325	1.379
238.375	3.447
238.425	0.475
238.475	3.931
238.525	11.732
238.575	20.719
238.625	22.306
238.675	43.706
238.725	36.490
238.775	23.923
238.825	31.545
238.875	16.330
238.925	42.133
238.975	30.764
239.025	25.693
239.075	19.543
239.125	24.601
239.175	37.268
239.225	6.428
239.275	13.608
239.325	25.529
239.375	23.967
239.425	42.822
239.475	26.438

Approved: July 06, 2016

SmartTune Wizard - Summary

```
Optimization Summary

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Start Time: 7/5/2016 9:09:35 AM
End Time: 7/5/2016 9:11:50 AM

Daily Performance Check - [Passed]  Optimum value(s): N/A
        Obtained Intensity (Be 9.0122): 5395.56
        Obtained Intensity (Mg 23.985): 243827.68
        Obtained Intensity (In 114.904): 70583.16
        Obtained Intensity (U 238.05): 142756.33
        Obtained Intensity (Bkgd 220): 0.20
        Obtained Formula (CeO 155.9 / Ce 139.905): 0.020 (=4055.54 / 204395.37)
        Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.011 (=2346.65 / 204395.37)
```

Report Date/Time: Tuesday, July 05, 2016 09:11:50
Page 1

Approved: July 06, 2016

SmartTune Wizard - Details

```
Optimization Details

SmartTune file: C:\NexIONData\Wizard\SmartTune\ESI SmartTune Fullmicrobac.swz

Optimization Status

Start Time: 7/5/2016 9:09:35 AM

Daily Performance Check
        Optimization Settings:
                Method: C:\NexIONData\Method\ESI Daily Performance.mth.
                Intensity Criterion: Be 9.0122 > 2000
                Intensity Criterion: Mg 23.985 > 15000
                Intensity Criterion: In 114.904 > 40000
                Intensity Criterion: U 238.05 > 30000
                Intensity Criterion: Bkgd 220 <= 5
                Formula Criterion: CeO 155.9 / Ce 139.905 <= 0.025
                Formula Criterion: Ce++ 69.9527 / Ce 139.905 <= 0.03

        Optimization Results:
        Initial Try
                Obtained Intensity (Be 9.0122): 5395.56
                Obtained Intensity (Mg 23.985): 243827.68
                Obtained Intensity (In 114.904): 70583.16
                Obtained Intensity (U 238.05): 142756.33
                Obtained Intensity (Bkgd 220): 0.20
                Obtained Formula (CeO 155.9 / Ce 139.905): 0.020 (=4055.54 / 204395.37)
                Obtained Formula (Ce++ 69.9527 / Ce 139.905): 0.011 (=2346.65 / 204395.37)
[Passed]  Optimum value(s): N/A

End Time: 7/5/2016 9:11:50 AM
```


Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: Blank
Sample Date/Time: Tuesday, July 05, 2016 09:23:52
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	59516.7	3.0				ug/L		Standard
	Be	9	10.0	50.0				ug/L		Standard
	Al	27	1871.8	6.0				ug/L		Standard
	Sc	45	26715.2	1.7				ug/L		Standard
	Ti	47	53.7	10.6				ug/L		Standard
	V	51	1419.3	5.4				ug/L		Standard
	Cr	52	7059.3	2.4				ug/L		Standard
	Cr	53	398.3	5.9				ug/L		Standard
	Mn	55	824.4	4.6				ug/L		Standard
	Co	59	166.7	12.5				ug/L		Standard
	Ni	60	391.3	3.3				ug/L		Standard
	Cu	65	147.0	7.8				ug/L		Standard
	Zn	66	183.7	12.0				ug/L		Standard
>	**Ge**	72	422997.6	2.0				ug/L		Standard
	As	75	-58.4	17.2				ug/L		Standard
	Se	82	14.9	7.9				ug/L		Standard
	Se-1	77	57.7	18.5				ug/L		Standard
>	**Ga**	71	113.3	5.1				mg/L		Standard
	Rb	85	26.7	21.7				ug/L		Standard
	Y	89	383780.2	1.3				ug/L		Standard
>	**Rh**	103	18.3	41.7				ug/L		Standard
	Mo	98	30.6	11.5				ug/L		Standard
	Ag	107	83.7	25.8				ug/L		Standard
	Cd	111	3.9	25.5				mg/L		Standard
	Cd	114	15.3	95.4				ug/L		Standard
>	**In**	115	515399.7	1.8				ug/L		Standard
	Sn	118	1043.4	21.4				ug/L		Standard
	Sb	123	226.5	22.8				ug/L		Standard
	Ba	135	30.7	16.1				ug/L		Standard
	Ce	140	120.0	28.9				ug/L		Standard
>	**Tb**	159	785817.3	2.8				ug/L		Standard
	Ho	165	6.7	86.6				ug/L		Standard
	Tl	203	5.7	53.9				ug/L		Standard
	Tl	205	8.3	69.3				ug/L		Standard
	Pb	206	200.7	7.0				ug/L		Standard
	Pb	207	176.7	10.5				ug/L		Standard
	Pb	208	721.7	1.4				ug/L		Standard
	U	238	7.7	74.2				ug/L		Standard
>	**Bi**	209	444562.8	2.6				ug/L		Standard

Sample ID: Blank
Report Date/Time: Tuesday, July 05, 2016 09:26:09
Page 1

Approved: July 06, 2016

	Na	23	0.0		mg/L	Standard
	Mg	24	38.3	32.8	mg/L	Standard
	K	39	33.3	77.0	mg/L	Standard
	Ca	43	48.3	26.0	mg/L	Standard
	Fe	54	130.9	15.4	mg/L	Standard
	Fe	57	153.3	25.3	mg/L	Standard
>	**Sc-1**	45	26715.2	1.7	mg/L	Standard
	Cl	35	102686.2	2.0	ug/L	Standard
	Kr	83	1.7	91.7	ug/L	Standard
	Br	81	1896.8	8.1	ug/L	Standard
	P	31	18491.3	3.9	ug/L	Standard
	S	34	3733.8	4.4	ug/L	Standard
	Sr	88	46.7	22.3	ug/L	Standard
	C	12	63.3	18.2	mg/L	Standard
	N	14	0.0		mg/L	Standard
	Hg	202	23.3	89.2	mg/L	Standard
	Dy	164	22.7	67.5	mg/L	Standard
	Ho-1	165	6.7	86.6	mg/L	Standard
	Er	166	13.3	43.3	mg/L	Standard
	I	127	12398.4	3.2	mg/L	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Blank
Report Date/Time: Tuesday, July 05, 2016 09:26:09
Page 2

Approved: July 06, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Blank
Report Date/Time: Tuesday, July 05, 2016 09:26:09
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: Standard 1
Sample Date/Time: Tuesday, July 05, 2016 09:27:04
Number of Replicates: 3
Autosampler Position: 1
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	59608.8	3.4				ug/L	59517	Standard
	Be	9	11.7	99.0				ug/L	10	Standard
	Al	27	1765.1	2.2				ug/L	1872	Standard
	Sc	45	26738.6	3.0				ug/L	26715	Standard
	Ti	47	56.0	10.9				ug/L	54	Standard
	V	51	1412.8	3.0				ug/L	1419	Standard
	Cr	52	6923.2	1.6				ug/L	7059	Standard
	Cr	53	398.3	10.8				ug/L	398	Standard
	Mn	55	831.4	1.9				ug/L	824	Standard
	Co	59	156.3	6.7				ug/L	167	Standard
	Ni	60	385.3	6.4				ug/L	391	Standard
	Cu	65	141.0	6.8				ug/L	147	Standard
	Zn	66	188.3	20.5				ug/L	184	Standard
>	**Ge**	72	430414.5	1.6				ug/L	422998	Standard
	As	75	-83.8	26.9				ug/L	-58	Standard
	Se	82	11.6	23.6				ug/L	15	Standard
	Se-1	77	58.3	8.8				ug/L	58	Standard
>	**Ga**	71	106.7	5.4				mg/L	113	Standard
	Rb	85	20.0	43.3				ug/L	27	Standard
	Y	89	380199.7	3.4				ug/L	383780	Standard
>	**Rh**	103	21.7	35.3				ug/L	18	Standard
	Mo	98	21.9	39.1				ug/L	31	Standard
	Ag	107	67.3	14.3				ug/L	84	Standard
	Cd	111	3.0	67.2				mg/L	4	Standard
	Cd	114	12.1	77.3				ug/L	15	Standard
>	**In**	115	507604.1	0.9				ug/L	515400	Standard
	Sn	118	1040.0	33.5				ug/L	1043	Standard
	Sb	123	109.1	24.9				ug/L	227	Standard
	Ba	135	34.0	11.8				ug/L	31	Standard
	Ce	140	93.3	17.2				ug/L	120	Standard
>	**Tb**	159	780753.8	1.6				ug/L	785817	Standard
	Ho	165	5.0	100.0				ug/L	7	Standard
	Tl	203	4.3	13.3				ug/L	6	Standard
	Tl	205	5.0	100.0				ug/L	8	Standard
	Pb	206	190.3	10.9				ug/L	201	Standard
	Pb	207	165.3	4.7				ug/L	177	Standard
	Pb	208	731.3	6.4				ug/L	722	Standard
	U	238	4.3	93.3				ug/L	8	Standard
>	**Bi**	209	442740.5	1.3				ug/L	444563	Standard

Sample ID: Standard 1
Report Date/Time: Tuesday, July 05, 2016 09:29:21
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	mg/L	0	Standard
Mg	24	41.7	36.7	mg/L	38	Standard
K	39	20.0	43.3	mg/L	33	Standard
Ca	43	45.0	22.2	mg/L	48	Standard
Fe	54	134.3	13.6	mg/L	131	Standard
Fe	57	140.0	27.9	mg/L	153	Standard
> **Sc-1**	45	26738.6	3.0	mg/L	26715	Standard
Cl	35	100953.2	1.4	ug/L	102686	Standard
Kr	83	3.0	33.3	ug/L	2	Standard
Br	81	1896.8	10.3	ug/L	1897	Standard
P	31	18573.1	4.4	ug/L	18491	Standard
S	34	3695.5	3.9	ug/L	3734	Standard
Sr	88	35.0	37.8	ug/L	47	Standard
C	12	80.0	33.1	mg/L	63	Standard
N	14	0.0		mg/L	0	Standard
Hg	202	33.3	45.8	mg/L	23	Standard
Dy	164	5.7	101.5	mg/L	23	Standard
Ho-1	165	5.0	100.0	mg/L	7	Standard
Er	166	20.0	50.0	mg/L	13	Standard
I	127	12588.6	2.6	mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72			
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: Standard 1
Report Date/Time: Tuesday, July 05, 2016 09:29:21
Page 2

Approved: July 06, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 1
Report Date/Time: Tuesday, July 05, 2016 09:29:21
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: Standard 2
Sample Date/Time: Tuesday, July 05, 2016 09:30:15
Number of Replicates: 3
Autosampler Position: 2
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	60323.2	1.6				ug/L	59517	Standard
	Be	9	71.7	28.2				ug/L	10	Standard
	Al	27	9626.4	1.5				ug/L	1872	Standard
	Sc	45	27469.9	2.0				ug/L	26715	Standard
	Ti	47	85.3	2.4				ug/L	54	Standard
	V	51	1714.1	1.7				ug/L	1419	Standard
	Cr	52	7550.9	1.7				ug/L	7059	Standard
	Cr	53	423.3	16.8				ug/L	398	Standard
	Mn	55	1185.4	1.3				ug/L	824	Standard
	Co	59	528.0	1.7				ug/L	167	Standard
	Ni	60	484.0	2.5				ug/L	391	Standard
	Cu	65	250.7	2.8				ug/L	147	Standard
	Zn	66	239.0	6.2				ug/L	184	Standard
>	Ge	72	437972.2	1.6				ug/L	422998	Standard
	As	75	-2.1	857.6				ug/L	-58	Standard
	Se	82	31.1	5.9				ug/L	15	Standard
	Se-1	77	56.3	13.4				ug/L	58	Standard
>	Ga	71	63.3	27.7				mg/L	113	Standard
	Rb	85	15.0	66.7				ug/L	27	Standard
	Y	89	387008.6	2.3				ug/L	383780	Standard
>	Rh	103	20.0	25.0				ug/L	18	Standard
	Mo	98	291.7	4.3				ug/L	31	Standard
	Ag	107	421.0	6.2				ug/L	84	Standard
	Cd	111	116.1	3.2				mg/L	4	Standard
	Cd	114	283.7	10.5				ug/L	15	Standard
>	In	115	519547.9	2.4				ug/L	515400	Standard
	Sn	118	1351.7	28.3				ug/L	1043	Standard
	Sb	123	306.2	6.2				ug/L	227	Standard
	Ba	135	149.3	0.4				ug/L	31	Standard
	Ce	140	90.0	34.7				ug/L	120	Standard
>	Tb	159	809068.9	3.1				ug/L	785817	Standard
	Ho	165	6.7	114.6				ug/L	7	Standard
	Tl	203	436.3	1.5				ug/L	6	Standard
	Tl	205	401.7	14.9				ug/L	8	Standard
	Pb	206	484.0	2.3				ug/L	201	Standard
	Pb	207	448.3	6.0				ug/L	177	Standard
	Pb	208	1813.0	1.7				ug/L	722	Standard
	U	238	366.7	3.7				ug/L	8	Standard
>	Bi	209	452738.0	3.1				ug/L	444563	Standard

Sample ID: Standard 2
Report Date/Time: Tuesday, July 05, 2016 09:32:32
Page 1

Approved: July 06, 2016

	Na	23	0.0		mg/L	0	Standard
	Mg	24	31.7	71.2	mg/L	38	Standard
	K	39	31.7	32.9	mg/L	33	Standard
	Ca	43	38.3	54.3	mg/L	48	Standard
	Fe	54	154.4	34.0	mg/L	131	Standard
	Fe	57	161.7	7.1	mg/L	153	Standard
>	**Sc-1**	45	27469.9	2.0	mg/L	26715	Standard
	Cl	35	102529.8	1.1	ug/L	102686	Standard
	Kr	83	2.0	86.6	ug/L	2	Standard
	Br	81	2006.8	8.1	ug/L	1897	Standard
	P	31	18007.4	3.0	ug/L	18491	Standard
	S	34	3722.1	4.1	ug/L	3734	Standard
	Sr	88	36.7	34.3	ug/L	47	Standard
	C	12	66.7	31.2	mg/L	63	Standard
	N	14	0.0		mg/L	0	Standard
	Hg	202	20.0	50.0	mg/L	23	Standard
	Dy	164	16.3	72.3	mg/L	23	Standard
	Ho-1	165	6.7	114.6	mg/L	7	Standard
	Er	166	6.7	86.6	mg/L	13	Standard
	I	127	12276.6	1.8	mg/L	12398	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72			
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: Standard 2
Report Date/Time: Tuesday, July 05, 2016 09:32:32
Page 2

Approved: July 06, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 2
Report Date/Time: Tuesday, July 05, 2016 09:32:32
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: Standard 3
Sample Date/Time: Tuesday, July 05, 2016 09:33:27
Number of Replicates: 3
Autosampler Position: 3
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	57281.5	3.4				ug/L	59517	Standard
	Be	9	42113.5	3.7	**50.0000**	0.890	1.8	ug/L	10	Standard
	Al	27	7411475.5	3.0	**50.0000**	0.930	1.9	ug/L	1872	Standard
	Sc	45	26479.8	2.2				ug/L	26715	Standard
	Ti	47	23890.5	1.1	**100.0000**	1.454	1.5	ug/L	54	Standard
	V	51	367062.7	0.8	**50.0000**	0.623	1.2	ug/L	1419	Standard
	Cr	52	367921.4	0.8	**50.0000**	0.612	1.2	ug/L	7059	Standard
	Cr	53	45347.8	2.4	**50.0000**	1.443	2.9	ug/L	398	Standard
	Mn	55	322217.3	2.0	**50.0000**	1.201	2.4	ug/L	824	Standard
	Co	59	345404.6	2.7	**50.0000**	1.575	3.2	ug/L	167	Standard
	Ni	60	95595.3	0.9	**50.0000**	0.634	1.3	ug/L	391	Standard
	Cu	65	95818.0	0.3	**50.0000**	0.238	0.5	ug/L	147	Standard
	Zn	66	56390.7	0.6	**50.0000**	0.243	0.5	ug/L	184	Standard
>	Ge	72	424214.4	0.4				ug/L	422998	Standard
	As	75	57453.1	0.7	**50.0000**	0.400	0.8	ug/L	-58	Standard
	Se	82	5987.9	1.8	**50.0000**	0.997	2.0	ug/L	15	Standard
	Se-1	77	3928.8	1.0	**50.0000**	0.491	1.0	ug/L	58	Standard
>	Ga	71	106.7	18.9				mg/L	113	Standard
	Rb	85	830.0	5.8				ug/L	27	Standard
	Y	89	376829.6	1.9				ug/L	383780	Standard
>	Rh	103	41.7	50.0				ug/L	18	Standard
	Mo	98	260891.0	0.5	**100.0000**	0.431	0.4	ug/L	31	Standard
	Ag	107	336697.0	1.4	**50.0000**	0.799	1.6	ug/L	84	Standard
	Cd	111	106142.8	0.9	**50.0000**	0.551	1.1	mg/L	4	Standard
	Cd	114	257891.1	2.0	**50.0000**	1.111	2.2	ug/L	15	Standard
>	In	115	505888.2	0.3				ug/L	515400	Standard
	Sn	118	283890.2	1.3	**50.0000**	0.797	1.6	ug/L	1043	Standard
	Sb	123	221275.8	2.3	**50.0000**	1.243	2.5	ug/L	227	Standard
	Ba	135	99917.1	1.9	**50.0000**	0.986	2.0	ug/L	31	Standard
	Ce	140	75.0	11.5				ug/L	120	Standard
>	Tb	159	787467.6	1.4				ug/L	785817	Standard
	Ho	165	11.7	65.5				ug/L	7	Standard
	Tl	203	403541.2	1.3	**50.0000**	0.368	0.7	ug/L	6	Standard
	Tl	205	348997.5	1.4	**50.0000**	0.411	0.8	ug/L	8	Standard
	Pb	206	249753.6	0.4	**50.0000**	0.480	1.0	ug/L	201	Standard
	Pb	207	226687.6	0.5	**50.0000**	0.605	1.2	ug/L	177	Standard
	Pb	208	907607.7	0.4	**50.0000**	0.234	0.5	ug/L	722	Standard
	U	238	342418.7	0.3	**50.0000**	0.445	0.9	ug/L	8	Standard
>	Bi	209	443560.9	0.8				ug/L	444563	Standard

Sample ID: Standard 3
Report Date/Time: Tuesday, July 05, 2016 09:35:44
Page 1

Approved: July 06, 2016

Na	23	0.0					mg/L	0	Standard
Mg	24	4067.2	1.5	**5.0000**	0.180	3.6	mg/L	38	Standard
K	39	1208.4	5.6	**5.0000**	0.216	4.3	mg/L	33	Standard
Ca	43	65.0	13.3	**5.0000**	1.694	33.9	mg/L	48	Standard
Fe	54	6344.5	2.9	**5.0000**	0.255	5.1	mg/L	131	Standard
Fe	57	1803.4	5.9	**5.0000**	0.400	8.0	mg/L	153	Standard
> **Sc-1**	45	26479.8	2.2				mg/L	26715	Standard
Cl	35	106266.8	1.7				ug/L	102686	Standard
Kr	83	3.0	0.0				ug/L	2	Standard
Br	81	1793.4	10.4				ug/L	1897	Standard
P	31	20133.4	2.6				ug/L	18491	Standard
S	34	4464.0	2.3				ug/L	3734	Standard
Sr	88	53.3	30.1				ug/L	47	Standard
C	12	56.7	10.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	23	Standard
Dy	164	3.0	201.4				mg/L	23	Standard
Ho-1	165	11.7	65.5				mg/L	7	Standard
Er	166	6.7	173.2				mg/L	13	Standard
I	127	20156.7	3.1				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72			
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: Standard 3
Report Date/Time: Tuesday, July 05, 2016 09:35:44
Page 2

Approved: July 06, 2016

⌊	Rb	85
⌈	Y	89
⌊>	Rh	103
⌈	Mo	98
|	Ag	107
|	Cd	111
|	Cd	114
|>	In	115
|	Sn	118
|	Sb	123
⌊	Ba	135
⌈	Ce	140
⌊>	Tb	159
⌈	Ho	165
|	Tl	203
|	Tl	205
|	Pb	206
|	Pb	207
|	Pb	208
|	U	238
⌊>	Bi	209
⌈	Na	23
|	Mg	24
|	K	39
|	Ca	43
|	Fe	54
|	Fe	57
⌊>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: Standard 3
Report Date/Time: Tuesday, July 05, 2016 09:35:44
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: Standard 4
Sample Date/Time: Tuesday, July 05, 2016 09:36:38
Number of Replicates: 3
Autosampler Position: 4
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	57177.7	2.3				ug/L	59517	Standard
	Be	9	83060.5	2.4	**99.4055**	0.082	0.1	ug/L	10	Standard
	Al	27	14427511.1	3.4	**98.7777**	5.601	5.7	ug/L	1872	Standard
	Sc	45	26401.3	1.0				ug/L	26715	Standard
	Ti	47	47521.1	2.0	**200.1084**	5.056	2.5	ug/L	54	Standard
	V	51	713200.7	1.1	**98.9035**	1.288	1.3	ug/L	1419	Standard
	Cr	52	707310.9	1.3	**98.7506**	1.789	1.8	ug/L	7059	Standard
	Cr	53	88258.2	3.7	**99.1029**	4.182	4.2	ug/L	398	Standard
	Mn	55	623978.6	0.6	**98.7110**	0.856	0.9	ug/L	824	Standard
	Co	59	668953.9	1.1	**98.6636**	0.987	1.0	ug/L	167	Standard
	Ni	60	184177.7	1.1	**98.4903**	1.424	1.4	ug/L	391	Standard
	Cu	65	184335.6	1.0	**98.3606**	1.396	1.4	ug/L	147	Standard
	Zn	66	108556.5	1.3	**98.4331**	1.767	1.8	ug/L	184	Standard
>	Ge	72	421989.5	0.5				ug/L	422998	Standard
	As	75	111881.9	1.6	**98.9050**	1.625	1.6	ug/L	-58	Standard
	Se	82	11551.8	0.8	**98.5578**	0.398	0.4	ug/L	15	Standard
	Se-1	77	7465.5	2.6	**98.0156**	2.561	2.6	ug/L	58	Standard
>	Ga	71	106.7	9.8				mg/L	113	Standard
	Rb	85	1656.8	1.9				ug/L	27	Standard
	Y	89	377045.5	0.7				ug/L	383780	Standard
>	Rh	103	60.0	30.0				ug/L	18	Standard
	Mo	98	516282.8	1.1	**200.0976**	4.353	2.2	ug/L	31	Standard
	Ag	107	642362.7	1.8	**98.2175**	1.938	2.0	ug/L	84	Standard
	Cd	111	203500.3	0.9	**98.4606**	1.356	1.4	mg/L	4	Standard
	Cd	114	493153.4	2.5	**98.3306**	2.664	2.7	ug/L	15	Standard
>	In	115	500187.8	1.5				ug/L	515400	Standard
	Sn	118	565406.4	0.3	**100.4534**	1.261	1.3	ug/L	1043	Standard
	Sb	123	451037.5	1.1	**101.5325**	2.258	2.2	ug/L	227	Standard
	Ba	135	192700.4	0.7	**98.7674**	1.578	1.6	ug/L	31	Standard
	Ce	140	151.7	19.0				ug/L	120	Standard
>	Tb	159	778055.5	0.3				ug/L	785817	Standard
	Ho	165	8.3	34.6				ug/L	7	Standard
	Tl	203	777662.3	0.5	**98.8652**	1.045	1.1	ug/L	6	Standard
	Tl	205	682449.5	1.4	**99.5981**	1.807	1.8	ug/L	8	Standard
	Pb	206	486716.7	1.0	**99.4428**	1.465	1.5	ug/L	201	Standard
	Pb	207	438279.6	1.0	**99.0466**	1.369	1.4	ug/L	177	Standard
	Pb	208	1766496.8	0.2	**99.3804**	0.506	0.5	ug/L	722	Standard
	U	238	675732.3	0.9	**100.0490**	1.332	1.3	ug/L	8	Standard
>	Bi	209	437245.8	0.6				ug/L	444563	Standard

Sample ID: Standard 4
Report Date/Time: Tuesday, July 05, 2016 09:38:55
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**19.9850**	34.606	173.2	mg/L	0	Standard
Mg	24	7703.6	2.1	**9.7561**	0.217	2.2	mg/L	38	Standard
K	39	2193.5	1.4	**9.5883**	0.205	2.1	mg/L	33	Standard
Ca	43	106.7	17.7	**11.0903**	2.878	26.0	mg/L	48	Standard
Fe	54	12288.8	3.6	**9.9066**	0.331	3.3	mg/L	131	Standard
Fe	57	3242.0	5.1	**9.6801**	0.435	4.5	mg/L	153	Standard
Sc-1	45	26401.3	1.0				mg/L	26715	Standard
Cl	35	104664.3	1.8				ug/L	102686	Standard
Kr	83	2.3	65.5				ug/L	2	Standard
Br	81	1870.1	15.3				ug/L	1897	Standard
P	31	20584.0	1.0				ug/L	18491	Standard
S	34	4714.1	1.1				ug/L	3734	Standard
Sr	88	46.7	34.4				ug/L	47	Standard
C	12	53.3	65.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	23	Standard
Dy	164	6.0	187.1				mg/L	23	Standard
Ho-1	165	8.3	34.6				mg/L	7	Standard
Er	166	13.3	86.6				mg/L	13	Standard
I	127	2513.5	5.2				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72			
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: Standard 4
Report Date/Time: Tuesday, July 05, 2016 09:38:55
Page 2

Approved: July 06, 2016

	Rb	85
	Y	89
>	Rh	103
	Mo	98
	Ag	107
	Cd	111
	Cd	114
>	In	115
	Sn	118
	Sb	123
	Ba	135
	Ce	140
>	Tb	159
	Ho	165
	Tl	203
	Tl	205
	Pb	206
	Pb	207
	Pb	208
	U	238
>	Bi	209
	Na	23
	Mg	24
	K	39
	Ca	43
	Fe	54
	Fe	57
>	Sc-1	45
	Cl	35
	Kr	83
	Br	81
	P	31
	S	34
	Sr	88
	C	12
	N	14
	Hg	202
	Dy	164
	Ho-1	165
	Er	166
	I	127

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Corr. Coef.	Na	23	Correlation coefficient < 0.998
Corr. Coef.	Ca	43	Correlation coefficient < 0.998

Sample ID: Standard 4
Report Date/Time: Tuesday, July 05, 2016 09:38:55
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 1
Sample Date/Time: Tuesday, July 05, 2016 09:39:52
Number of Replicates: 3
Autosampler Position: 201
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	53835.5	1.3				ug/L	59517	Standard
	Be	9	41558.6	3.6	**52.8023**	1.315	2.5	ug/L	10	Standard
	Al	27	7307731.0	1.7	**53.0856**	0.399	0.8	ug/L	1872	Standard
	Sc	45	25019.0	1.3				ug/L	26715	Standard
	Ti	47	23741.9	0.8	**105.4666**	0.518	0.5	ug/L	54	Standard
	V	51	361047.7	1.2	**52.7948**	0.251	0.5	ug/L	1419	Standard
	Cr	52	363213.6	1.6	**53.1053**	0.333	0.6	ug/L	7059	Standard
	Cr	53	44986.7	2.2	**53.1545**	1.109	2.1	ug/L	398	Standard
	Mn	55	313201.8	0.7	**52.2689**	0.422	0.8	ug/L	824	Standard
	Co	59	335791.2	0.4	**52.2971**	0.377	0.7	ug/L	167	Standard
	Ni	60	94287.8	0.7	**53.1623**	0.658	1.2	ug/L	391	Standard
	Cu	65	93679.1	0.8	**52.7602**	0.861	1.6	ug/L	147	Standard
	Zn	66	56044.1	0.6	**53.6021**	0.933	1.7	ug/L	184	Standard
>	Ge	72	399555.8	1.1				ug/L	422998	Standard
	As	75	55787.2	0.6	**52.1110**	0.279	0.5	ug/L	-58	Standard
	Se	82	5765.3	1.8	**51.8507**	0.353	0.7	ug/L	15	Standard
	Se-1	77	3764.5	1.9	**51.8829**	0.395	0.8	ug/L	58	Standard
>	Ga	71	83.3	6.9				mg/L	113	Standard
	Rb	85	805.0	6.3				ug/L	27	Standard
	Y	89	346814.9	1.4				ug/L	383780	Standard
>	Rh	103	33.3	45.8				ug/L	18	Standard
	Mo	98	255667.8	1.2	**103.8817**	3.429	3.3	ug/L	31	Standard
	Ag	107	326245.9	1.0	**52.2934**	1.513	2.9	ug/L	84	Standard
	Cd	111	104309.3	0.8	**52.9007**	0.886	1.7	mg/L	4	Standard
	Cd	114	252291.0	1.4	**52.7265**	0.923	1.8	ug/L	15	Standard
>	In	115	477243.7	2.4				ug/L	515400	Standard
	Sn	118	285175.5	1.0	**53.0339**	1.593	3.0	ug/L	1043	Standard
	Sb	123	227297.1	0.9	**53.6358**	1.767	3.3	ug/L	227	Standard
	Ba	135	97510.4	0.5	**52.3806**	1.184	2.3	ug/L	31	Standard
	Ce	140	70.0	25.8				ug/L	120	Standard
>	Tb	159	738225.1	1.8				ug/L	785817	Standard
	Ho	165	15.0	33.3				ug/L	7	Standard
	Tl	203	396474.4	1.1	**52.8617**	0.678	1.3	ug/L	6	Standard
	Tl	205	345053.4	1.1	**52.8086**	0.248	0.5	ug/L	8	Standard
	Pb	206	245442.3	1.4	**52.5746**	1.060	2.0	ug/L	201	Standard
	Pb	207	222933.5	0.8	**52.8178**	0.762	1.4	ug/L	177	Standard
	Pb	208	899249.4	0.4	**53.0368**	0.329	0.6	ug/L	722	Standard
	U	238	340452.5	0.4	**52.8663**	0.560	1.1	ug/L	8	Standard
>	Bi	209	416909.4	1.0				ug/L	444563	Standard

Sample ID: QC Std 1
Report Date/Time: Tuesday, July 05, 2016 09:42:09
Page 1

Approved: July 06, 2016

Na	23	3.3	86.6	**41.6409**	36.067	86.6	mg/L	0	Standard
Mg	24	4005.5	5.7	**5.3372**	0.289	5.4	mg/L	38	Standard
K	39	1045.0	2.9	**4.7561**	0.149	3.1	mg/L	33	Standard
Ca	43	70.0	14.3	**5.8755**	1.572	26.8	mg/L	48	Standard
Fe	54	6522.2	2.4	**5.4988**	0.200	3.6	mg/L	131	Standard
Fe	57	1745.1	3.0	**5.2936**	0.243	4.6	mg/L	153	Standard
> **Sc-1**	45	25019.0	1.3				mg/L	26715	Standard
Cl	35	101421.7	1.2				ug/L	102686	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	1796.8	11.2				ug/L	1897	Standard
P	31	20091.6	1.5				ug/L	18491	Standard
S	34	4492.3	1.8				ug/L	3734	Standard
Sr	88	45.0	19.2				ug/L	47	Standard
C	12	46.7	75.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	23	Standard
Dy	164	5.7	94.7				mg/L	23	Standard
Ho-1	165	15.0	33.3				mg/L	7	Standard
Er	166	20.0	86.6				mg/L	13	Standard
I	127	16966.2	1.4				mg/L	12398	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9	105.605		
	Al	27	106.171		
	Sc	45			
	Ti	47	105.467		
	V	51	105.590		
	Cr	52	106.211		
	Cr	53			
	Mn	55	104.538		
	Co	59	104.594		
	Ni	60	106.325		
	Cu	65	105.520		
	Zn	66	107.204		
>	Ge	72		94.458	
	As	75	104.222		
	Se	82	103.701		
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 1
Report Date/Time: Tuesday, July 05, 2016 09:42:09
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	103.882		
Ag	107	104.587		
Cd	111	105.801		
Cd	114			
> In	115		92.597	
Sn	118	106.068		
Sb	123	107.272		
Ba	135	104.761		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	105.723		
Tl	205			
Pb	206	105.149		
Pb	207	105.636		
Pb	208	106.074		
U	238	105.733		
> Bi	209		93.780	
Na	23	832.818		
Mg	24	106.744		
K	39	95.121		
Ca	43	117.511		
Fe	54	109.976		
Fe	57	105.873		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 1	Na	23	
QC Std 1	Ca	43	

Sample ID: QC Std 1
Report Date/Time: Tuesday, July 05, 2016 09:42:09
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 2
Sample Date/Time: Tuesday, July 05, 2016 09:43:05
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	56789.7	3.8				ug/L	59517	Standard
	Be	9	30.0	88.2	**0.0057**	0.034	592.0	ug/L	10	Standard
	Al	27	4072.2	25.4	**0.0156**	0.007	44.7	ug/L	1872	Standard
	Sc	45	26544.9	1.0				ug/L	26715	Standard
	Ti	47	56.0	5.4	**-0.0095**	0.012	131.0	ug/L	54	Standard
	V	51	1171.6	3.3	**-0.0159**	0.006	36.7	ug/L	1419	Standard
	Cr	52	6250.0	0.8	**-0.0905**	0.015	16.3	ug/L	7059	Standard
	Cr	53	385.0	13.7	**0.0257**	0.057	222.8	ug/L	398	Standard
	Mn	55	2770.3	4.9	**0.3100**	0.021	6.9	ug/L	824	Standard
	Co	59	278.3	16.3	**0.0162**	0.007	42.4	ug/L	167	Standard
	Ni	60	392.3	4.9	**0.0114**	0.011	92.8	ug/L	391	Standard
	Cu	65	169.0	8.4	**0.0117**	0.008	68.8	ug/L	147	Standard
	Zn	66	12934.5	0.1	**11.6450**	0.106	0.9	ug/L	184	Standard
>	**Ge**	72	419946.1	0.8				ug/L	422998	Standard
	As	75	-19.0	138.3	**0.0348**	0.023	67.0	ug/L	-58	Standard
	Se	82	20.1	22.5	**-0.0330**	0.040	122.3	ug/L	15	Standard
	Se-1	77	54.7	17.6	**0.0573**	0.124	216.2	ug/L	58	Standard
>	**Ga**	71	73.3	10.4				mg/L	113	Standard
	Rb	85	16.7	17.3				ug/L	27	Standard
	Y	89	372487.8	2.2				ug/L	383780	Standard
>	**Rh**	103	16.7	96.4				ug/L	18	Standard
	Mo	98	327.7	10.6	**0.1179**	0.015	12.4	ug/L	31	Standard
	Ag	107	169.7	22.8	**0.0139**	0.006	43.9	ug/L	84	Standard
	Cd	111	29.4	39.1	**0.0101**	0.006	55.8	mg/L	4	Standard
	Cd	114	75.2	27.2	**0.0104**	0.004	39.7	ug/L	15	Standard
>	**In**	115	501312.6	1.2				ug/L	515400	Standard
	Sn	118	1608.4	30.0	**0.1044**	0.084	80.4	ug/L	1043	Standard
	Sb	123	948.7	24.4	**0.1968**	0.052	26.4	ug/L	227	Standard
	Ba	135	97.7	19.3	**0.0263**	0.010	38.8	ug/L	31	Standard
	Ce	140	30.0	16.7				ug/L	120	Standard
>	**Tb**	159	766439.3	3.0				ug/L	785817	Standard
	Ho	165	1.7	173.2				ug/L	7	Standard
	Tl	203	130.3	29.7	**0.0128**	0.005	39.8	ug/L	6	Standard
	Tl	205	98.3	56.2	**0.0075**	0.008	109.1	ug/L	8	Standard
	Pb	206	365.3	8.6	**0.0283**	0.008	26.9	ug/L	201	Standard
	Pb	207	309.0	15.5	**0.0213**	0.012	55.8	ug/L	177	Standard
	Pb	208	1295.7	7.5	**0.0236**	0.007	28.3	ug/L	722	Standard
	U	238	108.0	28.7	**0.0134**	0.005	35.7	ug/L	8	Standard
>	**Bi**	209	442314.8	1.9				ug/L	444563	Standard

Sample ID: QC Std 2
Report Date/Time: Tuesday, July 05, 2016 09:45:22
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	23.3	53.9	**-0.0040**	0.016	402.6	mg/L	38	Standard
K	39	36.7	7.9	**0.0323**	0.013	39.7	mg/L	33	Standard
Ca	43	40.0	33.1	**0.4592**	2.104	458.2	mg/L	48	Standard
Fe	54	162.7	1.8	**0.0154**	0.001	9.5	mg/L	131	Standard
Fe	57	140.0	18.9	**-0.0460**	0.079	170.9	mg/L	153	Standard
> **Sc-1**	45	26544.9	1.0				mg/L	26715	Standard
Cl	35	101376.8	2.0				ug/L	102686	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1840.1	4.8				ug/L	1897	Standard
P	31	19113.7	2.8				ug/L	18491	Standard
S	34	4273.9	2.9				ug/L	3734	Standard
Sr	88	53.3	51.6				ug/L	47	Standard
C	12	53.3	21.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	23	Standard
Dy	164	6.3	93.2				mg/L	23	Standard
Ho-1	165	1.7	173.2				mg/L	7	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	11554.4	2.8				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.279	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 2
Report Date/Time: Tuesday, July 05, 2016 09:45:22
Page 2

Approved: July 06, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	97.267
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	99.494
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 2	Zn	66	
QC Std 2	Ca	43	

Sample ID: QC Std 2
Report Date/Time: Tuesday, July 05, 2016 09:45:22
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 3
Sample Date/Time: Tuesday, July 05, 2016 09:46:18
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	54322.3	4.4				ug/L	59517	Standard
	Be	9	135.0	16.1	**0.1399**	0.033	23.7	ug/L	10	Standard
	Al	27	1825.1	5.2	**0.0007**	0.001	96.2	ug/L	1872	Standard
	Sc	45	25576.6	3.2				ug/L	26715	Standard
	Ti	47	54.0		**-0.0134**	0.004	31.4	ug/L	54	Standard
	V	51	3677.6	2.6	**0.3442**	0.009	2.5	ug/L	1419	Standard
	Cr	52	11611.8	1.6	**0.7019**	0.005	0.7	ug/L	7059	Standard
	Cr	53	1015.0	5.3	**0.7630**	0.070	9.2	ug/L	398	Standard
	Mn	55	3542.4	1.4	**0.4443**	0.018	4.0	ug/L	824	Standard
	Co	59	2671.6	0.4	**0.3789**	0.007	1.9	ug/L	167	Standard
	Ni	60	3156.7	2.5	**1.5345**	0.073	4.7	ug/L	391	Standard
	Cu	65	1505.7	2.6	**0.7455**	0.036	4.8	ug/L	147	Standard
	Zn	66	7058.0	1.6	**6.4127**	0.220	3.4	ug/L	184	Standard
>	Ge	72	411736.9	1.8				ug/L	422998	Standard
	As	75	374.7	7.0	**0.3910**	0.024	6.1	ug/L	-58	Standard
	Se	82	56.6	11.5	**0.2895**	0.047	16.4	ug/L	15	Standard
	Se-1	77	76.7	8.7	**0.3701**	0.085	23.1	ug/L	58	Standard
>	Ga	71	63.3	31.9				mg/L	113	Standard
	Rb	85	23.3	53.9				ug/L	27	Standard
	Y	89	359086.1	1.7				ug/L	383780	Standard
>	Rh	103	23.3	32.7				ug/L	18	Standard
	Mo	98	119.6	23.9	**0.0392**	0.011	28.9	ug/L	31	Standard
	Ag	107	2559.5	2.0	**0.3941**	0.005	1.3	ug/L	84	Standard
	Cd	111	480.1	3.9	**0.2368**	0.004	1.9	mg/L	4	Standard
	Cd	114	1176.2	1.3	**0.2388**	0.002	0.9	ug/L	15	Standard
>	In	115	482062.0	2.1				ug/L	515400	Standard
	Sn	118	1013.4	35.3	**0.0066**	0.068	1039.3	ug/L	1043	Standard
	Sb	123	1750.5	1.7	**0.3926**	0.006	1.5	ug/L	227	Standard
	Ba	135	1380.7	4.6	**0.7107**	0.031	4.4	ug/L	31	Standard
	Ce	140	10.0					ug/L	120	Standard
>	Tb	159	750015.0	1.7				ug/L	785817	Standard
	Ho	165	3.3	86.6				ug/L	7	Standard
	Tl	203	633.0	3.0	**0.0778**	0.004	4.6	ug/L	6	Standard
	Tl	205	506.7	19.9	**0.0679**	0.014	20.8	ug/L	8	Standard
	Pb	206	1185.4	4.9	**0.1994**	0.013	6.7	ug/L	201	Standard
	Pb	207	1006.4	3.4	**0.1822**	0.011	6.3	ug/L	177	Standard
	Pb	208	4150.2	2.4	**0.1875**	0.002	0.9	ug/L	722	Standard
	U	238	2538.5	1.5	**0.3776**	0.011	3.0	ug/L	8	Standard
>	Bi	209	432540.0	1.8				ug/L	444563	Standard

Sample ID: QC Std 3
Report Date/Time: Tuesday, July 05, 2016 09:48:34
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**19.7989**	34.284	173.2	mg/L	0	Standard
Mg	24	25.0	34.6	**-0.0006**	0.012	1975.3	mg/L	38	Standard
K	39	28.3	10.2	**0.0001**	0.010	8097.2	mg/L	33	Standard
Ca	43	40.0	25.0	**0.6800**	1.556	228.8	mg/L	48	Standard
Fe	54	117.2	44.0	**-0.0169**	0.047	277.9	mg/L	131	Standard
Fe	57	171.7	7.3	**0.0745**	0.055	73.8	mg/L	153	Standard
> **Sc-1**	45	25576.6	3.2				mg/L	26715	Standard
Cl	35	98276.4	0.6				ug/L	102686	Standard
Kr	83	2.3	49.5				ug/L	2	Standard
Br	81	1690.1	6.4				ug/L	1897	Standard
P	31	13502.7	0.8				ug/L	18491	Standard
S	34	4085.6	1.8				ug/L	3734	Standard
Sr	88	66.7	26.3				ug/L	47	Standard
C	12	76.7	41.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	114.6				mg/L	23	Standard
Dy	164	5.7	202.4				mg/L	23	Standard
Ho-1	165	3.3	86.6				mg/L	7	Standard
Er	166	20.0	50.0				mg/L	13	Standard
I	127	1623.4	6.5				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	69.930		
Al	27	0.072		
Sc	45			
Ti	47			
V	51	86.043		
Cr	52	87.740		
Cr	53			
Mn	55	88.869		
Co	59	94.737		
Ni	60	95.908		
Cu	65	93.188		
Zn	66	102.603		
> Ge	72		97.338	
As	75	97.743		
Se	82	72.386		
Se-1	77			
> Ga	71			

Sample ID: QC Std 3
Report Date/Time: Tuesday, July 05, 2016 09:48:34
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	98.529		
Cd	111	98.678		
Cd	114			
> In	115			93.532
Sn	118			
Sb	123	98.139		
Ba	135	94.758		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	97.225		
Tl	205			
Pb	206			
Pb	207			
Pb	208	93.760		
U	238	94.395		
> Bi	209			97.296
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 3	Be	9	
QC Std 3	Al	27	

Sample ID: QC Std 3
Report Date/Time: Tuesday, July 05, 2016 09:48:34
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 1
Sample Date/Time: Tuesday, July 05, 2016 09:50:16
Number of Replicates: 3
Autosampler Position: 201
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	58169.8	2.1				ug/L	59517	Standard
	Be	9	41368.1	1.5	**48.6540**	0.517	1.1	ug/L	10	Standard
	Al	27	7290991.8	1.0	**49.0258**	0.636	1.3	ug/L	1872	Standard
	Sc	45	27236.1	0.9				ug/L	26715	Standard
	Ti	47	24092.5	1.0	**97.0294**	0.687	0.7	ug/L	54	Standard
	V	51	363336.2	1.3	**48.1642**	0.610	1.3	ug/L	1419	Standard
	Cr	52	362571.6	0.5	**47.9815**	0.112	0.2	ug/L	7059	Standard
	Cr	53	45230.8	1.6	**48.4251**	0.571	1.2	ug/L	398	Standard
	Mn	55	318203.0	1.3	**48.1419**	0.321	0.7	ug/L	824	Standard
	Co	59	342610.9	1.0	**48.3824**	0.165	0.3	ug/L	167	Standard
	Ni	60	93896.7	0.3	**47.9869**	0.223	0.5	ug/L	391	Standard
	Cu	65	94104.0	0.8	**48.0490**	0.164	0.3	ug/L	147	Standard
	Zn	66	55724.3	0.3	**48.3106**	0.453	0.9	ug/L	184	Standard
>	**Ge**	72	440608.2	0.7				ug/L	422998	Standard
	As	75	56127.2	1.3	**47.5458**	0.447	0.9	ug/L	-58	Standard
	Se	82	5821.4	0.5	**47.4627**	0.137	0.3	ug/L	15	Standard
	Se-1	77	3826.5	3.3	**47.7694**	1.288	2.7	ug/L	58	Standard
>	**Ga**	71	96.7	16.6				mg/L	113	Standard
	Rb	85	805.0	10.8				ug/L	27	Standard
	Y	89	385514.8	0.9				ug/L	383780	Standard
>	**Rh**	103	40.0	12.5				ug/L	18	Standard
	Mo	98	257542.0	0.9	**95.7910**	0.867	0.9	ug/L	31	Standard
	Ag	107	326830.8	0.5	**47.9581**	0.259	0.5	ug/L	84	Standard
	Cd	111	103201.3	0.9	**47.9228**	0.449	0.9	mg/L	4	Standard
	Cd	114	256440.4	0.5	**49.0734**	0.209	0.4	ug/L	15	Standard
>	**In**	115	521082.8	0.1				ug/L	515400	Standard
	Sn	118	282756.0	1.5	**48.1210**	0.666	1.4	ug/L	1043	Standard
	Sb	123	227361.0	1.0	**49.1107**	0.551	1.1	ug/L	227	Standard
	Ba	135	97385.0	0.8	**47.8931**	0.346	0.7	ug/L	31	Standard
	Ce	140	93.3	13.5				ug/L	120	Standard
>	**Tb**	159	801308.4	1.6				ug/L	785817	Standard
	Ho	165	5.0	100.0				ug/L	7	Standard
	Tl	203	395362.2	0.5	**48.0233**	0.556	1.2	ug/L	6	Standard
	Tl	205	344540.9	1.9	**48.0428**	1.206	2.5	ug/L	8	Standard
	Pb	206	246479.1	0.9	**48.0951**	0.873	1.8	ug/L	201	Standard
	Pb	207	223686.9	1.1	**48.2810**	1.143	2.4	ug/L	177	Standard
	Pb	208	894445.5	0.5	**48.0575**	0.742	1.5	ug/L	722	Standard
	U	238	335748.7	0.7	**47.4954**	0.465	1.0	ug/L	8	Standard
>	**Bi**	209	457654.3	1.5				ug/L	444563	Standard

Sample ID: QC Std 1
Report Date/Time: Tuesday, July 05, 2016 09:52:33
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3785.5	3.0	**4.6294**	0.144	3.1	mg/L	38	Standard
K	39	1143.4	3.2	**4.7796**	0.115	2.4	mg/L	33	Standard
Ca	43	73.3	10.4	**5.4487**	1.278	23.5	mg/L	48	Standard
Fe	54	6420.9	3.1	**4.9595**	0.123	2.5	mg/L	131	Standard
Fe	57	1733.4	5.4	**4.7880**	0.327	6.8	mg/L	153	Standard
Sc-1	45	27236.1	0.9				mg/L	26715	Standard
Cl	35	104987.2	0.9				ug/L	102686	Standard
Kr	83	2.7	78.1				ug/L	2	Standard
Br	81	1720.1	6.9				ug/L	1897	Standard
P	31	20764.2	1.5				ug/L	18491	Standard
S	34	4507.3	5.1				ug/L	3734	Standard
Sr	88	68.3	15.2				ug/L	47	Standard
C	12	60.0	33.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	86.6				mg/L	23	Standard
Dy	164	2.7	204.8				mg/L	23	Standard
Ho-1	165	5.0	100.0				mg/L	7	Standard
Er	166	13.3	86.6				mg/L	13	Standard
I	127	18129.2	1.8				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	97.308		
Al	27	98.052		
Sc	45			
Ti	47	97.029		
V	51	96.328		
Cr	52	95.963		
Cr	53			
Mn	55	96.284		
Co	59	96.765		
Ni	60	95.974		
Cu	65	96.098		
Zn	66	96.621		
Ge	72		104.163	
As	75	95.092		
Se	82	94.925		
Se-1	77			
Ga	71			

Sample ID: QC Std 1
Report Date/Time: Tuesday, July 05, 2016 09:52:33
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	95.791		
Ag	107	95.916		
Cd	111	95.846		
Cd	114			
> In	115		101.103	
Sn	118	96.242		
Sb	123	98.221		
Ba	135	95.786		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	96.047		
Tl	205			
Pb	206	96.190		
Pb	207	96.562		
Pb	208	96.115		
U	238	94.991		
> Bi	209		102.945	
Na	23	0.100		
Mg	24	92.588		
K	39	95.592		
Ca	43	108.974		
Fe	54	99.191		
Fe	57	95.761		
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 1	Na	23	
QC Std 1	Ca	43	

Sample ID: QC Std 1
Report Date/Time: Tuesday, July 05, 2016 09:52:33
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 2
Sample Date/Time: Tuesday, July 05, 2016 09:53:29
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	59412.9	1.2				ug/L	59517	Standard
	Be	9	20.0	25.0	-0.0081	0.006	68.9	ug/L	10	Standard
	Al	27	2121.8	7.7	0.0015	0.001	60.2	ug/L	1872	Standard
	Sc	45	28268.0	1.9				ug/L	26715	Standard
	Ti	47	48.3	20.7	-0.0579	0.042	71.8	ug/L	54	Standard
	V	51	1136.5	4.4	-0.0333	0.007	20.2	ug/L	1419	Standard
	Cr	52	6308.6	1.2	-0.1530	0.019	12.3	ug/L	7059	Standard
	Cr	53	343.3	8.0	-0.0523	0.026	48.8	ug/L	398	Standard
	Mn	55	813.0	4.5	-0.0118	0.005	39.4	ug/L	824	Standard
	Co	59	237.3	8.5	0.0073	0.003	35.7	ug/L	167	Standard
	Ni	60	355.7	3.7	-0.0236	0.005	21.0	ug/L	391	Standard
	Cu	65	128.0	8.9	-0.0157	0.007	42.7	ug/L	147	Standard
	Zn	66	246.7	1.0	0.0480	0.006	11.5	ug/L	184	Standard
>	Ge	72	456136.8	1.7				ug/L	422998	Standard
	As	75	-59.5	37.3	0.0031	0.018	584.9	ug/L	-58	Standard
	Se	82	16.8	17.9	-0.0733	0.024	32.6	ug/L	15	Standard
	Se-1	77	52.7	26.9	-0.0225	0.185	821.1	ug/L	58	Standard
>	Ga	71	86.7	16.7				mg/L	113	Standard
	Rb	85	21.7	26.6				ug/L	27	Standard
	Y	89	400244.6	1.1				ug/L	383780	Standard
>	Rh	103	33.3	31.2				ug/L	18	Standard
	Mo	98	228.4	8.4	0.0733	0.009	11.9	ug/L	31	Standard
	Ag	107	125.0	9.2	0.0057	0.001	22.2	ug/L	84	Standard
	Cd	111	13.3	49.0	0.0018	0.003	158.0	mg/L	4	Standard
	Cd	114	43.3	20.3	0.0035	0.002	53.2	ug/L	15	Standard
>	In	115	539308.9	3.0				ug/L	515400	Standard
	Sn	118	1368.4	33.9	0.0448	0.074	166.2	ug/L	1043	Standard
	Sb	123	874.6	30.5	0.1667	0.057	34.0	ug/L	227	Standard
	Ba	135	48.0	24.0	-0.0010	0.005	512.8	ug/L	31	Standard
	Ce	140	23.3	24.7				ug/L	120	Standard
>	Tb	159	818055.2	1.1				ug/L	785817	Standard
	Ho	165	3.3	173.2				ug/L	7	Standard
	Tl	203	69.0	24.1	0.0045	0.002	42.7	ug/L	6	Standard
	Tl	205	78.3	3.7	0.0038	0.000	10.6	ug/L	8	Standard
	Pb	206	261.0	2.0	0.0038	0.001	27.1	ug/L	201	Standard
	Pb	207	226.0	8.6	-0.0006	0.004	641.6	ug/L	177	Standard
	Pb	208	908.7	7.8	-0.0013	0.003	268.8	ug/L	722	Standard
	U	238	59.7	23.5	0.0057	0.002	33.6	ug/L	8	Standard
>	Bi	209	472449.6	0.6				ug/L	444563	Standard

Sample ID: QC Std 2
Report Date/Time: Tuesday, July 05, 2016 09:55:46
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	48.3	41.8	**0.0240**	0.025	103.6	mg/L	38	Standard
K	39	30.0	44.1	**-0.0053**	0.055	1036.6	mg/L	33	Standard
Ca	43	26.7	39.0	**-1.8930**	1.609	85.0	mg/L	48	Standard
Fe	54	145.8	1.7	**-0.0055**	0.000	4.8	mg/L	131	Standard
Fe	57	115.0	11.5	**-0.1455**	0.035	23.7	mg/L	153	Standard
Sc-1	45	28268.0	1.9				mg/L	26715	Standard
Cl	35	105093.2	1.1				ug/L	102686	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	1796.8	3.9				ug/L	1897	Standard
P	31	20867.7	1.2				ug/L	18491	Standard
S	34	4170.6	0.8				ug/L	3734	Standard
Sr	88	55.0	47.2				ug/L	47	Standard
C	12	50.0	40.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	23	Standard
Dy	164	6.3	173.2				mg/L	23	Standard
Ho-1	165	3.3	173.2				mg/L	7	Standard
Er	166	6.7	173.2				mg/L	13	Standard
I	127	12480.1	2.6				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		107.834	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 2
Report Date/Time: Tuesday, July 05, 2016 09:55:46
Page 2

Approved: July 06, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	104.639
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	106.273
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 2	Ca	43	
QC Std 2	Fe	57	

Sample ID: QC Std 2
Report Date/Time: Tuesday, July 05, 2016 09:55:46
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 3
Sample Date/Time: Tuesday, July 05, 2016 09:56:42
Number of Replicates: 3
Autosampler Position: 202
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	57425.3	1.0				ug/L	59517	Standard
	Be	9	171.7	14.4	**0.1737**	0.031	17.8	ug/L	10	Standard
	Al	27	1671.8	5.4	**-0.0010**	0.001	48.6	ug/L	1872	Standard
	Sc	45	27112.6	0.1				ug/L	26715	Standard
	Ti	47	37.3	13.7	**-0.0946**	0.023	24.4	ug/L	54	Standard
	V	51	3759.7	2.7	**0.3249**	0.017	5.4	ug/L	1419	Standard
	Cr	52	11667.8	2.0	**0.6122**	0.040	6.6	ug/L	7059	Standard
	Cr	53	990.0	3.0	**0.6672**	0.029	4.4	ug/L	398	Standard
	Mn	55	3647.8	3.0	**0.4270**	0.022	5.2	ug/L	824	Standard
	Co	59	2727.9	0.6	**0.3635**	0.006	1.7	ug/L	167	Standard
	Ni	60	3080.3	0.9	**1.3935**	0.015	1.1	ug/L	391	Standard
	Cu	65	1549.1	0.7	**0.7196**	0.012	1.6	ug/L	147	Standard
	Zn	66	7437.5	2.3	**6.3635**	0.253	4.0	ug/L	184	Standard
>	**Ge**	72	437157.7	1.7				ug/L	422998	Standard
	As	75	339.2	5.7	**0.3409**	0.013	3.8	ug/L	-58	Standard
	Se	82	54.1	10.9	**0.2403**	0.041	17.2	ug/L	15	Standard
	Se-1	77	89.3	5.2	**0.4718**	0.063	13.4	ug/L	58	Standard
>	**Ga**	71	75.0	50.3				mg/L	113	Standard
	Rb	85	26.7	47.2				ug/L	27	Standard
	Y	89	385694.7	1.5				ug/L	383780	Standard
>	**Rh**	103	21.7	13.3				ug/L	18	Standard
	Mo	98	78.3	0.7	**0.0204**	0.000	1.9	ug/L	31	Standard
	Ag	107	2513.9	3.9	**0.3598**	0.013	3.6	ug/L	84	Standard
	Cd	111	481.5	2.2	**0.2212**	0.006	2.8	mg/L	4	Standard
	Cd	114	1238.9	8.8	**0.2343**	0.020	8.3	ug/L	15	Standard
>	**In**	115	517133.7	0.6				ug/L	515400	Standard
	Sn	118	976.7	49.7	**-0.0132**	0.082	623.9	ug/L	1043	Standard
	Sb	123	1749.4	2.7	**0.3645**	0.008	2.2	ug/L	227	Standard
	Ba	135	1392.1	0.8	**0.6665**	0.005	0.7	ug/L	31	Standard
	Ce	140	8.3	34.6				ug/L	120	Standard
>	**Tb**	159	792688.6	1.9				ug/L	785817	Standard
	Ho	165	6.7	173.2				ug/L	7	Standard
	Tl	203	607.3	6.1	**0.0696**	0.005	7.2	ug/L	6	Standard
	Tl	205	523.3	13.0	**0.0656**	0.009	13.3	ug/L	8	Standard
	Pb	206	1216.0	1.0	**0.1901**	0.001	0.3	ug/L	201	Standard
	Pb	207	968.4	6.5	**0.1597**	0.013	8.4	ug/L	177	Standard
	Pb	208	4236.9	1.2	**0.1776**	0.005	3.1	ug/L	722	Standard
	U	238	2521.2	2.7	**0.3516**	0.013	3.6	ug/L	8	Standard
>	**Bi**	209	461086.3	1.2				ug/L	444563	Standard

Sample ID: QC Std 3
Report Date/Time: Tuesday, July 05, 2016 09:58:59
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	28.3	20.4	**0.0014**	0.007	496.5	mg/L	38	Standard
K	39	25.0	40.0	**-0.0214**	0.043	200.9	mg/L	33	Standard
Ca	43	43.3	17.6	**0.8409**	1.190	141.6	mg/L	48	Standard
Fe	54	94.1	29.6	**-0.0419**	0.022	52.6	mg/L	131	Standard
Fe	57	151.7	13.3	**-0.0190**	0.061	320.9	mg/L	153	Standard
> **Sc-1**	45	27112.6	0.1				mg/L	26715	Standard
Cl	35	101496.2	0.7				ug/L	102686	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1850.1	6.5				ug/L	1897	Standard
P	31	14078.2	0.9				ug/L	18491	Standard
S	34	4153.9	5.8				ug/L	3734	Standard
Sr	88	56.7	27.0				ug/L	47	Standard
C	12	53.3	43.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	0.0					mg/L	23	Standard
Dy	164	6.3	99.7				mg/L	23	Standard
Ho-1	165	6.7	173.2				mg/L	7	Standard
Er	166	6.7	173.2				mg/L	13	Standard
I	127	1555.1	12.5				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	86.835		
Al	27	-0.105		
Sc	45			
Ti	47			
V	51	81.222		
Cr	52	76.528		
Cr	53			
Mn	55	85.397		
Co	59	90.866		
Ni	60	87.096		
Cu	65	89.951		
Zn	66	101.816		
> Ge	72		103.348	
As	75	85.222		
Se	82	60.067		
Se-1	77			
> Ga	71			

Sample ID: QC Std 3
Report Date/Time: Tuesday, July 05, 2016 09:58:59
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98			
Ag	107	89.940		
Cd	111	92.173		
Cd	114			
> In	115			100.336
Sn	118			
Sb	123	91.126		
Ba	135	88.861		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	87.048		
Tl	205			
Pb	206			
Pb	207			
Pb	208	88.795		
U	238	87.897		
> Bi	209			103.717
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 3	Al	27	
QC Std 3	Se	82	

Sample ID: QC Std 3
Report Date/Time: Tuesday, July 05, 2016 09:58:59
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 4
Sample Date/Time: Tuesday, July 05, 2016 09:59:58
Number of Replicates: 3
Autosampler Position: 203
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	58373.9	2.3				ug/L	59517	Standard
	Be	9	16.7	34.6	**-0.0116**	0.007	57.1	ug/L	10	Standard
	Al	27	7244731.4	2.0	**48.5385**	0.251	0.5	ug/L	1872	Standard
	Sc	45	27491.6	2.6				ug/L	26715	Standard
	Ti	47	26462.4	1.5	**106.6713**	1.834	1.7	ug/L	54	Standard
	V	51	1512.6	18.0	**0.0220**	0.037	166.8	ug/L	1419	Standard
	Cr	52	7779.3	1.8	**0.0752**	0.021	28.0	ug/L	7059	Standard
	Cr	53	3347.0	3.9	**3.2061**	0.142	4.4	ug/L	398	Standard
	Mn	55	2353.2	59.7	**0.2265**	0.214	94.5	ug/L	824	Standard
	Co	59	391.7	3.0	**0.0303**	0.002	5.2	ug/L	167	Standard
	Ni	60	1062.7	4.3	**0.3458**	0.022	6.4	ug/L	391	Standard
	Cu	65	758.4	1.9	**0.3091**	0.008	2.6	ug/L	147	Standard
	Zn	66	1411.1	5.3	**1.0687**	0.063	5.9	ug/L	184	Standard
>	**Ge**	72	440319.5	0.2				ug/L	422998	Standard
	As	75	-70.0	30.8	**-0.0077**	0.018	238.3	ug/L	-58	Standard
	Se	82	13.4	27.6	**-0.0963**	0.031	31.8	ug/L	15	Standard
	Se-1	77	311.3	6.7	**3.2753**	0.257	7.8	ug/L	58	Standard
>	**Ga**	71	173.3	16.9				mg/L	113	Standard
	Rb	85	1570.1	11.6				ug/L	27	Standard
	Y	89	383720.2	1.1				ug/L	383780	Standard
>	**Rh**	103	15.0	57.7				ug/L	18	Standard
	Mo	98	259304.7	1.5	**96.9732**	2.021	2.1	ug/L	31	Standard
	Ag	107	298.3	7.6	**0.0320**	0.004	11.1	ug/L	84	Standard
	Cd	111	-198.8	8.2	**-0.0970**	0.008	8.4	mg/L	4	Standard
	Cd	114	476.2	6.6	**0.0871**	0.006	6.8	ug/L	15	Standard
>	**In**	115	518299.5	0.7				ug/L	515400	Standard
	Sn	118	1570.1	18.5	**0.0885**	0.048	54.2	ug/L	1043	Standard
	Sb	123	688.1	37.0	**0.1329**	0.054	40.8	ug/L	227	Standard
	Ba	135	1511.7	5.2	**0.7243**	0.044	6.1	ug/L	31	Standard
	Ce	140	441.7	2.4				ug/L	120	Standard
>	**Tb**	159	798601.3	0.6				ug/L	785817	Standard
	Ho	165	18.3	68.6				ug/L	7	Standard
	Tl	203	124.7	7.6	**0.0119**	0.001	10.4	ug/L	6	Standard
	Tl	205	96.7	29.9	**0.0070**	0.004	60.2	ug/L	8	Standard
	Pb	206	820.0	2.3	**0.1182**	0.005	4.3	ug/L	201	Standard
	Pb	207	695.3	3.3	**0.1057**	0.006	5.6	ug/L	177	Standard
	Pb	208	2917.4	0.3	**0.1118**	0.001	0.9	ug/L	722	Standard
	U	238	38.0	26.7	**0.0030**	0.001	47.3	ug/L	8	Standard
>	**Bi**	209	447566.5	0.9				ug/L	444563	Standard

Sample ID: QC Std 4
Report Date/Time: Tuesday, July 05, 2016 10:02:15
Page 1

Approved: July 06, 2016

Na	23	5.0	100.0	**57.7254**	58.680	101.7	mg/L	0	Standard
Mg	24	9931.6	2.3	**12.0868**	0.062	0.5	mg/L	38	Standard
K	39	1073.4	8.0	**4.4334**	0.273	6.2	mg/L	33	Standard
Ca	43	115.0	28.5	**11.7193**	4.981	42.5	mg/L	48	Standard
Fe	54	15793.8	2.5	**12.2551**	0.022	0.2	mg/L	131	Standard
Fe	57	4364.0	4.8	**12.6536**	0.356	2.8	mg/L	153	Standard
Sc-1	45	27491.6	2.6				mg/L	26715	Standard
Cl	35	106721.4	2.5				ug/L	102686	Standard
Kr	83	1.7	69.3				ug/L	2	Standard
Br	81	2000.1	10.4				ug/L	1897	Standard
P	31	19956.5	2.7				ug/L	18491	Standard
S	34	3973.9	6.5				ug/L	3734	Standard
Sr	88	75.0	13.3				ug/L	47	Standard
C	12	136.7	4.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	80.0	0.0				mg/L	23	Standard
Dy	164	14.8	39.1				mg/L	23	Standard
Ho-1	165	18.3	68.6				mg/L	7	Standard
Er	166	40.0					mg/L	13	Standard
I	127	2396.9	6.0				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27	0.971		
Sc	45			
Ti	47	106.671		
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		104.095	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 4
Report Date/Time: Tuesday, July 05, 2016 10:02:15
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	96.973		
Ag	107			
Cd	111			
Cd	114			
In	115			100.563
Sn	118			
Sb	123			
Ba	135			
Ce	140			
Tb	159			
Ho	165			
Tl	203			
Tl	205			
Pb	206			
Pb	207			
Pb	208			
U	238			
Bi	209			100.676
Na	23	461.803		
Mg	24	241.737		
K	39	88.667		
Ca	43	78.129		
Fe	54	98.041		
Fe	57	101.229		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 4	Al	27	
QC Std 4	Ba	135	
QC Std 4	Na	23	

Sample ID: QC Std 4
Report Date/Time: Tuesday, July 05, 2016 10:02:15
Page 3

Approved: July 06, 2016

QC Std 4	Mg	24
QC Std 4	Ca	43

Sample ID: QC Std 4
Report Date/Time: Tuesday, July 05, 2016 10:02:15
Page 4

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 5
Sample Date/Time: Tuesday, July 05, 2016 10:03:17
Number of Replicates: 3
Autosampler Position: 204
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	58993.0	1.3				ug/L	59517	Standard
	Be	9	85001.2	1.9	**98.6019**	1.685	1.7	ug/L	10	Standard
	Al	27	6321358.0	0.5	**41.9099**	0.712	1.7	ug/L	1872	Standard
	Sc	45	26818.7	3.0				ug/L	26715	Standard
	Ti	47	22236.0	1.7	**88.9928**	1.339	1.5	ug/L	54	Standard
	V	51	726317.3	0.4	**95.8788**	0.488	0.5	ug/L	1419	Standard
	Cr	52	712141.9	0.8	**94.6051**	0.860	0.9	ug/L	7059	Standard
	Cr	53	90801.3	1.1	**97.0401**	1.065	1.1	ug/L	398	Standard
	Mn	55	638561.6	0.4	**96.1618**	0.252	0.3	ug/L	824	Standard
	Co	59	676060.4	0.2	**94.9228**	0.023	0.0	ug/L	167	Standard
	Ni	60	185772.6	0.1	**94.5619**	0.177	0.2	ug/L	391	Standard
	Cu	65	186681.3	0.3	**94.8225**	0.204	0.2	ug/L	147	Standard
	Zn	66	113558.3	0.9	**98.0179**	0.939	1.0	ug/L	184	Standard
>	Ge	72	443273.1	0.2				ug/L	422998	Standard
	As	75	116279.2	0.3	**97.8554**	0.264	0.3	ug/L	-58	Standard
	Se	82	11822.1	0.5	**96.0160**	0.269	0.3	ug/L	15	Standard
	Se-1	77	8046.5	0.6	**100.5862**	0.536	0.5	ug/L	58	Standard
>	Ga	71	176.7	16.3				mg/L	113	Standard
	Rb	85	743.4	11.5				ug/L	27	Standard
	Y	89	387252.6	1.9				ug/L	383780	Standard
>	Rh	103	36.7	43.8				ug/L	18	Standard
	Mo	98	221408.4	1.2	**82.1531**	1.082	1.3	ug/L	31	Standard
	Ag	107	569943.8	0.4	**83.4398**	0.317	0.4	ug/L	84	Standard
	Cd	111	208059.9	0.8	**96.3891**	1.070	1.1	mg/L	4	Standard
	Cd	114	504841.5	1.1	**96.3794**	0.834	0.9	ug/L	15	Standard
>	In	115	522337.6	0.3				ug/L	515400	Standard
	Sn	118	1488.4	20.9	**0.0727**	0.053	72.5	ug/L	1043	Standard
	Sb	123	445991.9	2.4	**96.1170**	2.148	2.2	ug/L	227	Standard
	Ba	135	192476.2	0.4	**94.4542**	0.161	0.2	ug/L	31	Standard
	Ce	140	63.3	29.9				ug/L	120	Standard
>	Tb	159	807206.6	0.7				ug/L	785817	Standard
	Ho	165	13.3	21.7				ug/L	7	Standard
	Tl	203	799136.9	1.1	**95.7679**	0.378	0.4	ug/L	6	Standard
	Tl	205	701248.1	0.9	**96.4717**	0.388	0.4	ug/L	8	Standard
	Pb	206	509858.8	0.7	**98.2043**	1.552	1.6	ug/L	201	Standard
	Pb	207	446077.7	0.5	**95.0325**	1.252	1.3	ug/L	177	Standard
	Pb	208	1809334.4	0.4	**95.9604**	1.210	1.3	ug/L	722	Standard
	U	238	675199.0	0.6	**94.2439**	1.346	1.4	ug/L	8	Standard
>	Bi	209	463829.2	0.9				ug/L	444563	Standard

Sample ID: QC Std 5
Report Date/Time: Tuesday, July 05, 2016 10:05:34
Page 1

Approved: July 06, 2016

Na	23	8.3	34.6	**97.7411**	35.296	36.1	mg/L	0	Standard
Mg	24	8068.8	2.4	**10.0676**	0.439	4.4	mg/L	38	Standard
K	39	913.4	6.3	**3.8609**	0.369	9.6	mg/L	33	Standard
Ca	43	135.0	3.7	**15.3007**	1.074	7.0	mg/L	48	Standard
Fe	54	13362.2	3.2	**10.6210**	0.522	4.9	mg/L	131	Standard
Fe	57	3627.1	4.9	**10.7240**	0.749	7.0	mg/L	153	Standard
> **Sc-1**	45	26818.7	3.0				mg/L	26715	Standard
Cl	35	104227.4	2.6				ug/L	102686	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	1820.1	4.5				ug/L	1897	Standard
P	31	17890.6	5.0				ug/L	18491	Standard
S	34	3797.1	2.3				ug/L	3734	Standard
Sr	88	56.7	31.0				ug/L	47	Standard
C	12	103.3	40.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0					mg/L	23	Standard
Dy	164	16.3	36.2				mg/L	23	Standard
Ho-1	165	13.3	21.7				mg/L	7	Standard
Er	166	6.7	86.6				mg/L	13	Standard
I	127	9746.5	1.1				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	98.602		
Al	27	0.838		
Sc	45			
Ti	47	88.993		
V	51	95.879		
Cr	52	94.605		
Cr	53			
Mn	55	96.162		
Co	59	94.923		
Ni	60	94.562		
Cu	65	94.823		
Zn	66	98.018		
> Ge	72		104.793	
As	75	97.855		
Se	82	96.016		
Se-1	77			
> Ga	71			

Sample ID: QC Std 5
Report Date/Time: Tuesday, July 05, 2016 10:05:34
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	82.153		
Ag	107	83.440		
Cd	111	96.389		
Cd	114			
In	115		101.346	
Sn	118			
Sb	123	96.117		
Ba	135	94.454		
Ce	140			
Tb	159			
Ho	165			
Tl	203	95.768		
Tl	205			
Pb	206			
Pb	207			
Pb	208	95.960		
U	238	94.244		
Bi	209		104.334	
Na	23	781.929		
Mg	24	201.352		
K	39	77.218		
Ca	43	102.004		
Fe	54	84.968		
Fe	57	85.792		
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 5	Al	27	
QC Std 5	Na	23	
QC Std 5	Mg	24	

Sample ID: QC Std 5
Report Date/Time: Tuesday, July 05, 2016 10:05:34
Page 3

Approved: July 06, 2016

QC Std 5	K	39

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, July 05, 2016 10:06:30
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	59831.3	2.1				ug/L	59517	Standard
	Be	9	43026.0	1.6	**49.2010**	0.807	1.6	ug/L	10	Standard
	Al	27	7352971.0	1.7	**48.0862**	1.810	3.8	ug/L	1872	Standard
	Sc	45	26534.9	2.7				ug/L	26715	Standard
	Ti	47	23836.7	1.1	**98.5966**	2.873	2.9	ug/L	54	Standard
	V	51	365784.8	1.3	**49.8014**	1.375	2.8	ug/L	1419	Standard
	Cr	52	363613.1	1.2	**49.4443**	1.281	2.6	ug/L	7059	Standard
	Cr	53	45628.6	1.6	**50.1882**	1.671	3.3	ug/L	398	Standard
	Mn	55	316216.3	1.1	**49.1365**	1.378	2.8	ug/L	824	Standard
	Co	59	343483.3	1.1	**49.8130**	1.198	2.4	ug/L	167	Standard
	Ni	60	94484.0	0.6	**49.5915**	1.019	2.1	ug/L	391	Standard
	Cu	65	93736.1	0.4	**49.1473**	0.729	1.5	ug/L	147	Standard
	Zn	66	55817.0	1.0	**49.6877**	0.390	0.8	ug/L	184	Standard
>	**Ge**	72	429169.2	1.8				ug/L	422998	Standard
	As	75	55939.1	0.5	**48.6605**	1.008	2.1	ug/L	-58	Standard
	Se	82	5820.2	1.0	**48.7343**	1.077	2.2	ug/L	15	Standard
	Se-1	77	3847.5	1.4	**49.3499**	1.185	2.4	ug/L	58	Standard
>	**Ga**	71	81.7	19.7				mg/L	113	Standard
	Rb	85	806.7	1.3				ug/L	27	Standard
	Y	89	373583.7	1.6				ug/L	383780	Standard
>	**Rh**	103	21.7	35.3				ug/L	18	Standard
	Mo	98	257231.3	1.0	**97.8031**	2.633	2.7	ug/L	31	Standard
	Ag	107	329273.0	1.1	**49.3919**	1.374	2.8	ug/L	84	Standard
	Cd	111	104357.1	0.5	**49.5334**	1.014	2.0	mg/L	4	Standard
	Cd	114	256663.0	0.5	**50.2059**	1.112	2.2	ug/L	15	Standard
>	**In**	115	509906.4	1.8				ug/L	515400	Standard
	Sn	118	281156.4	0.7	**48.9072**	0.530	1.1	ug/L	1043	Standard
	Sb	123	235448.6	1.7	**51.9870**	1.471	2.8	ug/L	227	Standard
	Ba	135	97479.2	0.7	**49.0045**	1.165	2.4	ug/L	31	Standard
	Ce	140	83.3	25.0				ug/L	120	Standard
>	**Tb**	159	786134.9	1.0				ug/L	785817	Standard
	Ho	165	11.7	24.7				ug/L	7	Standard
	Tl	203	401970.1	0.5	**50.0255**	0.983	2.0	ug/L	6	Standard
	Tl	205	351161.4	1.1	**50.1638**	0.887	1.8	ug/L	8	Standard
	Pb	206	252778.5	0.6	**50.5367**	1.059	2.1	ug/L	201	Standard
	Pb	207	228467.0	0.4	**50.5201**	0.932	1.8	ug/L	177	Standard
	Pb	208	912205.7	0.8	**50.2139**	0.866	1.7	ug/L	722	Standard
	U	238	344223.3	0.4	**49.8915**	0.966	1.9	ug/L	8	Standard
>	**Bi**	209	446729.0	1.8				ug/L	444563	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 10:08:46
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**19.5061**	33.777	173.2	mg/L	0	Standard
Mg	24	3785.5	6.4	**4.7554**	0.343	7.2	mg/L	38	Standard
K	39	1055.0	7.5	**4.5194**	0.293	6.5	mg/L	33	Standard
Ca	43	73.3	14.2	**5.7764**	1.991	34.5	mg/L	48	Standard
Fe	54	6345.3	4.0	**5.0384**	0.324	6.4	mg/L	131	Standard
Fe	57	1775.1	7.3	**5.0532**	0.317	6.3	mg/L	153	Standard
> **Sc-1**	45	26534.9	2.7				mg/L	26715	Standard
Cl	35	105006.0	1.0				ug/L	102686	Standard
Kr	83	3.0	57.7				ug/L	2	Standard
Br	81	1946.8	18.0				ug/L	1897	Standard
P	31	19901.4	4.7				ug/L	18491	Standard
S	34	4455.7	5.1				ug/L	3734	Standard
Sr	88	60.0	28.9				ug/L	47	Standard
C	12	70.0	28.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	57.3				mg/L	23	Standard
Dy	164	8.9	8.2				mg/L	23	Standard
Ho-1	165	11.7	24.7				mg/L	7	Standard
Er	166	23.3	65.5				mg/L	13	Standard
I	127	17641.9	1.6				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	98.402		
Al	27	96.172		
Sc	45			
Ti	47	98.597		
V	51	99.603		
Cr	52	98.889		
Cr	53			
Mn	55	98.273		
Co	59	99.626		
Ni	60	99.183		
Cu	65	98.295		
Zn	66	99.375		
> Ge	72		101.459	
As	75	97.321		
Se	82	97.469		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 10:08:46
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	97.803		
Ag	107	98.784		
Cd	111	99.067		
Cd	114			
In	115			98.934
Sn	118	97.814		
Sb	123	103.974		
Ba	135	98.009		
Ce	140			
Tb	159			
Ho	165			
Tl	203	100.051		
Tl	205			
Pb	206			
Pb	207			
Pb	208	100.428		
U	238	99.783		
Bi	209			100.487
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 10:08:46
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, July 05, 2016 10:09:41
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	62071.9	1.7				ug/L	59517	Standard
	Be	9	35.0	51.5	**0.0075**	0.020	267.5	ug/L	10	Standard
	Al	27	2741.9	25.6	**0.0048**	0.004	85.5	ug/L	1872	Standard
	Sc	45	27962.5	1.7				ug/L	26715	Standard
	Ti	47	42.3	20.6	**-0.0795**	0.031	39.5	ug/L	54	Standard
	V	51	1176.3	5.4	**-0.0258**	0.011	44.1	ug/L	1419	Standard
	Cr	52	6381.0	4.6	**-0.1306**	0.051	39.1	ug/L	7059	Standard
	Cr	53	451.7	12.2	**0.0676**	0.052	77.5	ug/L	398	Standard
	Mn	55	787.0	8.0	**-0.0137**	0.012	85.3	ug/L	824	Standard
	Co	59	275.7	10.8	**0.0132**	0.005	37.0	ug/L	167	Standard
	Ni	60	330.7	7.5	**-0.0333**	0.015	45.2	ug/L	391	Standard
	Cu	65	144.0	10.8	**-0.0068**	0.008	120.9	ug/L	147	Standard
	Zn	66	236.7	2.7	**0.0425**	0.003	6.1	ug/L	184	Standard
>	Ge	72	449380.2	2.1				ug/L	422998	Standard
	As	75	-11.5	135.8	**0.0423**	0.013	30.5	ug/L	-58	Standard
	Se	82	19.1	0.9	**-0.0525**	0.004	7.6	ug/L	15	Standard
	Se-1	77	63.7	8.7	**0.1212**	0.053	43.4	ug/L	58	Standard
>	Ga	71	75.0	17.6				mg/L	113	Standard
	Rb	85	21.7	48.0				ug/L	27	Standard
	Y	89	395093.7	1.1				ug/L	383780	Standard
>	Rh	103	18.3	15.7				ug/L	18	Standard
	Mo	98	300.7	7.3	**0.1000**	0.009	8.7	ug/L	31	Standard
	Ag	107	169.0	33.1	**0.0122**	0.008	67.3	ug/L	84	Standard
	Cd	111	19.2	19.5	**0.0045**	0.002	38.8	mg/L	4	Standard
	Cd	114	50.2	53.0	**0.0048**	0.005	104.8	ug/L	15	Standard
>	In	115	535175.7	0.7				ug/L	515400	Standard
	Sn	118	1420.1	24.7	**0.0550**	0.057	103.3	ug/L	1043	Standard
	Sb	123	1609.3	30.2	**0.3217**	0.100	31.0	ug/L	227	Standard
	Ba	135	59.3	37.5	**0.0048**	0.011	229.7	ug/L	31	Standard
	Ce	140	16.7	34.6				ug/L	120	Standard
>	Tb	159	823938.4	1.2				ug/L	785817	Standard
	Ho	165	3.3	86.6				ug/L	7	Standard
	Tl	203	100.0	27.0	**0.0082**	0.003	40.2	ug/L	6	Standard
	Tl	205	80.0	33.1	**0.0040**	0.004	89.7	ug/L	8	Standard
	Pb	206	276.7	12.2	**0.0068**	0.007	100.8	ug/L	201	Standard
	Pb	207	246.3	11.5	**0.0037**	0.006	173.1	ug/L	177	Standard
	Pb	208	996.0	14.6	**0.0033**	0.008	242.0	ug/L	722	Standard
	U	238	87.3	28.4	**0.0095**	0.004	37.1	ug/L	8	Standard
>	Bi	209	472387.3	0.9				ug/L	444563	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 10:11:58
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	36.7	64.4	**0.0106**	0.029	269.8	mg/L	38	Standard
K	39	26.7	65.8	**-0.0169**	0.075	445.3	mg/L	33	Standard
Ca	43	41.7	38.6	**0.4058**	2.515	619.8	mg/L	48	Standard
Fe	54	174.3	5.6	**0.0177**	0.009	50.0	mg/L	131	Standard
Fe	57	146.7	18.8	**-0.0473**	0.086	181.6	mg/L	153	Standard
> **Sc-1**	45	27962.5	1.7				mg/L	26715	Standard
Cl	35	108546.2	2.2				ug/L	102686	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	1846.8	4.6				ug/L	1897	Standard
P	31	19873.0	3.3				ug/L	18491	Standard
S	34	4410.6	6.9				ug/L	3734	Standard
Sr	88	48.3	26.0				ug/L	47	Standard
C	12	56.7	27.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	23	Standard
Dy	164	13.2	44.9				mg/L	23	Standard
Ho-1	165	3.3	86.6				mg/L	7	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	12096.5	1.7				mg/L	12398	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6			
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		106.237	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 10:11:58
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.837
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.259
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Sb	123	

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 10:11:58
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: PBS H4 WG574663-02
Sample Date/Time: Tuesday, July 05, 2016 10:15:30
Number of Replicates: 3
Autosampler Position: 205
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	65491.6	6.6				ug/L	59517	Standard
	Be	9	23.3	61.9	**-0.0073**	0.013	179.3	ug/L	10	Standard
	Al	27	9262.8	2.6	**0.0430**	0.002	5.0	ug/L	1872	Standard
	Sc	45	24905.5	0.4				ug/L	26715	Standard
	Ti	47	53.3	7.8	**-0.0067**	0.012	180.4	ug/L	54	Standard
	V	51	11580.0	21.4	**1.5451**	0.402	26.0	ug/L	1419	Standard
	Cr	52	18550.7	4.4	**1.8221**	0.199	10.9	ug/L	7059	Standard
	Cr	53	97195.7	8.9	**116.9078**	13.389	11.5	ug/L	398	Standard
	Mn	55	1234.7	4.6	**0.0783**	0.015	19.1	ug/L	824	Standard
	Co	59	246.7	10.0	**0.0138**	0.003	20.5	ug/L	167	Standard
	Ni	60	434.0	10.2	**0.0483**	0.018	37.4	ug/L	391	Standard
	Cu	65	264.0	13.4	**0.0713**	0.016	22.6	ug/L	147	Standard
	Zn	66	1357.1	2.0	**1.1577**	0.014	1.2	ug/L	184	Standard
>	Ge	72	394957.5	2.8				ug/L	422998	Standard
	As	75	-546.0	81.6	**-0.4703**	0.431	91.6	ug/L	-58	Standard
	Se	82	15.8	32.0	**-0.0604**	0.050	82.1	ug/L	15	Standard
	Se-1	77	8179.9	4.9	**115.0215**	8.790	7.6	ug/L	58	Standard
>	Ga	71	223.3	20.1				mg/L	113	Standard
	Rb	85	35.0	49.5				ug/L	27	Standard
	Y	89	345203.4	1.0				ug/L	383780	Standard
>	Rh	103	43.3	24.0				ug/L	18	Standard
	Mo	98	105.6	9.9	**0.0361**	0.005	12.5	ug/L	31	Standard
	Ag	107	267.0	48.1	**0.0327**	0.021	63.5	ug/L	84	Standard
	Cd	111	26.4	86.0	**0.0098**	0.012	120.1	mg/L	4	Standard
	Cd	114	74.6	104.8	**0.0116**	0.017	144.1	ug/L	15	Standard
>	In	115	454743.1	1.5				ug/L	515400	Standard
	Sn	118	1131.7	22.3	**0.0409**	0.052	125.9	ug/L	1043	Standard
	Sb	123	654.9	20.0	**0.1461**	0.034	23.5	ug/L	227	Standard
	Ba	135	108.3	26.2	**0.0372**	0.015	40.4	ug/L	31	Standard
	Ce	140	33.3	31.2				ug/L	120	Standard
>	Tb	159	780570.0	1.6				ug/L	785817	Standard
	Ho	165	6.7	43.3				ug/L	7	Standard
	Tl	203	134.0	57.5	**0.0133**	0.009	69.9	ug/L	6	Standard
	Tl	205	101.7	22.7	**0.0080**	0.003	40.1	ug/L	8	Standard
	Pb	206	393.0	10.4	**0.0347**	0.006	18.5	ug/L	201	Standard
	Pb	207	345.0	7.6	**0.0301**	0.004	13.7	ug/L	177	Standard
	Pb	208	1343.4	6.1	**0.0270**	0.003	10.4	ug/L	722	Standard
	U	238	18.7	52.6	**0.0003**	0.001	518.1	ug/L	8	Standard
>	Bi	209	437101.4	2.3				ug/L	444563	Standard

Sample ID: PBS H4 WG574663-02
Report Date/Time: Tuesday, July 05, 2016 10:17:46
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**20.8779**	36.153	173.2	mg/L	0	Standard
Mg	24	110.0	27.6	**0.1144**	0.040	35.4	mg/L	38	Standard
K	39	45.0	40.1	**0.0823**	0.086	104.1	mg/L	33	Standard
Ca	43	33.3	17.3	**-0.2499**	0.998	399.5	mg/L	48	Standard
Fe	54	118.6	27.5	**-0.0140**	0.029	204.1	mg/L	131	Standard
Fe	57	198.3	7.3	**0.1771**	0.046	26.0	mg/L	153	Standard
> **Sc-1**	45	24905.5	0.4				mg/L	26715	Standard
Cl	35	97454.4	0.7				ug/L	102686	Standard
Kr	83	2.3	99.0				ug/L	2	Standard
Br	81	1743.4	11.3				ug/L	1897	Standard
P	31	13639.5	3.1				ug/L	18491	Standard
S	34	4730.7	2.7				ug/L	3734	Standard
Sr	88	56.7	43.5				ug/L	47	Standard
C	12	106.7	42.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	23	Standard
Dy	164	22.7	48.4				mg/L	23	Standard
Ho-1	165	6.7	43.3				mg/L	7	Standard
Er	166	13.3	86.6				mg/L	13	Standard
I	127	28009.2	1.2				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.039	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		93.371	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBS H4 WG574663-02
Report Date/Time: Tuesday, July 05, 2016 10:17:46
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	88.231	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	98.322	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Cr 53 Upper, S, EEE	Cr	53	
As 75 Lower	As	75	
Se-1 77 Upper, S, EEE	Se-1	77	

Sample ID: PBS H4 WG574663-02
Report Date/Time: Tuesday, July 05, 2016 10:17:46
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: LCSS H4 WG574663-03
Sample Date/Time: Tuesday, July 05, 2016 10:18:41
Number of Replicates: 3
Autosampler Position: 206
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	89097.9	0.4				ug/L	59517	Standard
	Be	9	24382.9	0.9	**18.7020**	0.205	1.1	ug/L	10	Standard
	Al	27	19215.5	6.6	**0.0719**	0.005	7.4	ug/L	1872	Standard
	Sc	45	28077.7	3.6				ug/L	26715	Standard
	Ti	47	145.3	57.1	**0.3700**	0.343	92.8	ug/L	54	Standard
	V	51	212579.4	2.3	**29.6977**	1.062	3.6	ug/L	1419	Standard
	Cr	52	232740.8	1.5	**32.2112**	0.384	1.2	ug/L	7059	Standard
	Cr	53	248774.5	1.9	**283.2488**	5.159	1.8	ug/L	398	Standard
	Mn	55	174578.1	1.0	**27.8390**	0.323	1.2	ug/L	824	Standard
	Co	59	176300.5	0.7	**26.2811**	0.174	0.7	ug/L	167	Standard
	Ni	60	50092.7	0.4	**26.9485**	0.274	1.0	ug/L	391	Standard
	Cu	65	49677.3	1.4	**26.7504**	0.101	0.4	ug/L	147	Standard
	Zn	66	29338.0	0.9	**26.7882**	0.120	0.4	ug/L	184	Standard
>	**Ge**	72	417245.0	1.3				ug/L	422998	Standard
	As	75	27983.7	1.6	**25.0585**	0.389	1.6	ug/L	-58	Standard
	Se	82	2946.4	0.9	**25.2742**	0.435	1.7	ug/L	15	Standard
	Se-1	77	22278.7	1.2	**297.1877**	3.709	1.2	ug/L	58	Standard
>	**Ga**	71	496.7	11.8				mg/L	113	Standard
	Rb	85	156.7	9.2				ug/L	27	Standard
	Y	89	374825.2	1.4				ug/L	383780	Standard
>	**Rh**	103	63.3	39.7				ug/L	18	Standard
	Mo	98	1114.0	2.7	**0.4346**	0.005	1.1	ug/L	31	Standard
	Ag	107	170515.1	0.3	**26.7795**	0.381	1.4	ug/L	84	Standard
	Cd	111	51574.1	1.2	**25.6346**	0.385	1.5	mg/L	4	Standard
	Cd	114	125467.5	1.3	**25.6992**	0.368	1.4	ug/L	15	Standard
>	**In**	115	486841.8	1.7				ug/L	515400	Standard
	Sn	118	1626.8	27.1	**0.1157**	0.076	65.3	ug/L	1043	Standard
	Sb	123	109667.4	0.5	**25.3502**	0.343	1.4	ug/L	227	Standard
	Ba	135	52113.8	0.4	**27.4265**	0.446	1.6	ug/L	31	Standard
	Ce	140	126.7	16.4				ug/L	120	Standard
>	**Tb**	159	858869.4	0.8				ug/L	785817	Standard
	Ho	165	6.7	114.6				ug/L	7	Standard
	Tl	203	211401.4	1.2	**24.7568**	0.524	2.1	ug/L	6	Standard
	Tl	205	170864.7	0.8	**22.9661**	0.319	1.4	ug/L	8	Standard
	Pb	206	143557.2	0.8	**26.9866**	0.422	1.6	ug/L	201	Standard
	Pb	207	123747.0	1.0	**25.7254**	0.260	1.0	ug/L	177	Standard
	Pb	208	456532.1	0.8	**23.6244**	0.409	1.7	ug/L	722	Standard
	U	238	172264.4	0.6	**23.4936**	0.257	1.1	ug/L	8	Standard
>	**Bi**	209	474657.3	0.9				ug/L	444563	Standard

Sample ID: LCSS H4 WG574663-03
Report Date/Time: Tuesday, July 05, 2016 10:20:58
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	180.0	21.0	**0.1822**	0.050	27.7	mg/L	38	Standard
K	39	40.0	33.1	**0.0364**	0.051	141.0	mg/L	33	Standard
Ca	43	40.0	25.0	**0.1510**	1.720	1138.4	mg/L	48	Standard
Fe	54	176.5	29.4	**0.0183**	0.038	205.6	mg/L	131	Standard
Fe	57	171.7	8.9	**0.0243**	0.045	185.8	mg/L	153	Standard
Sc-1	45	28077.7	3.6				mg/L	26715	Standard
Cl	35	107579.9	0.5				ug/L	102686	Standard
Kr	83	2.3	99.0				ug/L	2	Standard
Br	81	1966.8	7.4				ug/L	1897	Standard
P	31	24441.4	1.4				ug/L	18491	Standard
S	34	5606.0	4.1				ug/L	3734	Standard
Sr	88	85.0	15.6				ug/L	47	Standard
C	12	93.3	52.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	13.3	86.6				mg/L	23	Standard
Dy	164	25.4	55.4				mg/L	23	Standard
Ho-1	165	6.7	114.6				mg/L	7	Standard
Er	166	26.7	94.4				mg/L	13	Standard
I	127	15756.5	2.7				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		149.702	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.640	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSS H4 WG574663-03
Report Date/Time: Tuesday, July 05, 2016 10:20:58
Page 2

Approved: July 06, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	94.459
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	106.769
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Cr 53 Upper, S, EEE	Cr	53	
Se-1 77 Upper, S, EEE	Se-1	77	

Sample ID: LCSS H4 WG574663-03
Report Date/Time: Tuesday, July 05, 2016 10:20:58
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145106 WG574663-01
Sample Date/Time: Tuesday, July 05, 2016 10:21:52
Number of Replicates: 3
Autosampler Position: 207
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	84418.1	3.8				ug/L	59517	Standard
	Be	9	1016.7	0.6	**0.7941**	0.034	4.3	ug/L	10	Standard
	Al	27	3563578.5	0.9	**16.5127**	0.485	2.9	ug/L	1872	Standard
	Sc	45	35351.2	2.6				ug/L	26715	Standard
	Ti	47	21288.6	2.1	**83.6919**	1.223	1.5	ug/L	54	Standard
	V	51	514875.3	0.7	**66.7274**	1.014	1.5	ug/L	1419	Standard
	Cr	52	156424.1	1.5	**19.6513**	0.300	1.5	ug/L	7059	Standard
	Cr	53	138593.1	3.1	**145.7297**	4.567	3.1	ug/L	398	Standard
	Mn	55	1832238.2	1.2	**271.3443**	4.602	1.7	ug/L	824	Standard
	Co	59	25524.2	1.7	**3.4972**	0.081	2.3	ug/L	167	Standard
	Ni	60	42103.8	1.2	**20.9006**	0.103	0.5	ug/L	391	Standard
	Cu	65	102546.6	1.4	**51.1390**	0.631	1.2	ug/L	147	Standard
	Zn	66	89291.0	1.4	**75.6836**	0.517	0.7	ug/L	184	Standard
>	Ge	72	451173.5	0.8				ug/L	422998	Standard
	As	75	16334.4	0.4	**13.5506**	0.097	0.7	ug/L	-58	Standard
	Se	82	503.0	2.3	**3.8161**	0.087	2.3	ug/L	15	Standard
	Se-1	77	10945.0	5.9	**134.6658**	8.375	6.2	ug/L	58	Standard
>	Ga	71	11064.0	1.9				mg/L	113	Standard
	Rb	85	36604.2	2.2				ug/L	27	Standard
	Y	89	859178.3	2.1				ug/L	383780	Standard
>	Rh	103	75.0	11.5				ug/L	18	Standard
	Mo	98	11321.5	2.3	**4.2188**	0.152	3.6	ug/L	31	Standard
	Ag	107	875.0	4.4	**0.1169**	0.006	4.9	ug/L	84	Standard
	Cd	111	4512.3	2.4	**2.0988**	0.038	1.8	mg/L	4	Standard
	Cd	114	11115.0	3.9	**2.1311**	0.107	5.0	ug/L	15	Standard
>	In	115	519235.1	1.6				ug/L	515400	Standard
	Sn	118	6929.9	4.0	**1.0075**	0.060	5.9	ug/L	1043	Standard
	Sb	123	4873.8	13.1	**1.0406**	0.140	13.5	ug/L	227	Standard
	Ba	135	276651.4	2.3	**136.6046**	3.722	2.7	ug/L	31	Standard
	Ce	140	1457468.2	2.4				ug/L	120	Standard
>	Tb	159	896312.5	0.8				ug/L	785817	Standard
	Ho	165	34457.5	1.6				ug/L	7	Standard
	Tl	203	1627.4	1.2	**0.1812**	0.001	0.3	ug/L	6	Standard
	Tl	205	1371.7	7.8	**0.1721**	0.016	9.1	ug/L	8	Standard
	Pb	206	182170.6	0.3	**33.2179**	0.218	0.7	ug/L	201	Standard
	Pb	207	150468.5	0.3	**30.3435**	0.324	1.1	ug/L	177	Standard
	Pb	208	590890.8	1.2	**29.6618**	0.348	1.2	ug/L	722	Standard
	U	238	21543.6	1.1	**2.8472**	0.058	2.0	ug/L	8	Standard
>	Bi	209	489471.3	0.9				ug/L	444563	Standard

Sample ID: L1606145106 WG574663-01
Report Date/Time: Tuesday, July 05, 2016 10:24:09
Page 1

Approved: July 06, 2016

Na	23	3.3	86.6	**29.9651**	25.947	86.6	mg/L	0	Standard
Mg	24	515.0	3.5	**0.4556**	0.029	6.3	mg/L	38	Standard
K	39	155.0	11.6	**0.3827**	0.046	11.9	mg/L	33	Standard
Ca	43	225.0	7.7	**20.9177**	2.599	12.4	mg/L	48	Standard
Fe	54	21358.0	2.6	**12.9039**	0.625	4.8	mg/L	131	Standard
Fe	57	5671.1	3.2	**12.8045**	0.603	4.7	mg/L	153	Standard
> **Sc-1**	45	35351.2	2.6				mg/L	26715	Standard
Cl	35	120813.8	1.3				ug/L	102686	Standard
Kr	83	1.0	0.0				ug/L	2	Standard
Br	81	5834.5	10.0				ug/L	1897	Standard
P	31	21410.1	2.6				ug/L	18491	Standard
S	34	5474.3	2.0				ug/L	3734	Standard
Sr	88	166.7	9.2				ug/L	47	Standard
C	12	1330.1	1.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	456.7	16.1				mg/L	23	Standard
Dy	164	52163.9	2.2				mg/L	23	Standard
Ho-1	165	34457.5	1.6				mg/L	7	Standard
Er	166	33219.7	1.7				mg/L	13	Standard
I	127	71483.4	4.2				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		141.839	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		106.661	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145106 WG574663-01
Report Date/Time: Tuesday, July 05, 2016 10:24:09
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		100.744
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.102
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Cr 53 Upper, S, EEE	Cr	53	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606145106 WG574663-01
Report Date/Time: Tuesday, July 05, 2016 10:24:09
Page 3

Approved: July 06, 2016

Se-1 77 Upper, S, EEE	Se-1	77
Ba 135 Upper, S, EEE	Ba	135

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145109S WG574663-04
Sample Date/Time: Tuesday, July 05, 2016 10:25:04
Number of Replicates: 3
Autosampler Position: 208
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	91812.3	0.3				ug/L	59517	Standard
	Be	9	25788.6	3.5	19.1948	0.629	3.3	ug/L	10	Standard
	Al	27	3689674.2	1.8	15.7082	0.304	1.9	ug/L	1872	Standard
	Sc	45	36654.3	1.2				ug/L	26715	Standard
	Ti	47	24746.9	1.9	95.2339	5.236	5.5	ug/L	54	Standard
	V	51	715478.9	2.3	90.7635	4.466	4.9	ug/L	1419	Standard
	Cr	52	364867.0	2.4	46.1080	2.802	6.1	ug/L	7059	Standard
	Cr	53	188822.7	1.2	194.3702	9.132	4.7	ug/L	398	Standard
	Mn	55	1803183.7	0.9	261.2009	11.281	4.3	ug/L	824	Standard
	Co	59	203800.3	1.3	27.4874	1.345	4.9	ug/L	167	Standard
	Ni	60	94872.8	0.3	46.3061	1.796	3.9	ug/L	391	Standard
	Cu	65	156908.3	1.1	76.5505	1.942	2.5	ug/L	147	Standard
	Zn	66	122522.4	0.6	101.6191	3.306	3.3	ug/L	184	Standard
>	Ge	72	461699.8	3.6				ug/L	422998	Standard
	As	75	44548.3	1.1	36.0611	1.478	4.1	ug/L	-58	Standard
	Se	82	3806.4	2.0	29.5786	1.649	5.6	ug/L	15	Standard
	Se-1	77	15638.4	1.1	188.4429	7.560	4.0	ug/L	58	Standard
>	Ga	71	12923.8	3.5				mg/L	113	Standard
	Rb	85	41130.8	2.5				ug/L	27	Standard
	Y	89	883461.3	1.6				ug/L	383780	Standard
>	Rh	103	81.7	36.9				ug/L	18	Standard
	Mo	98	11119.1	1.0	4.0936	0.201	4.9	ug/L	31	Standard
	Ag	107	169437.6	1.6	24.6587	1.344	5.5	ug/L	84	Standard
	Cd	111	57906.2	2.1	26.6733	1.497	5.6	mg/L	4	Standard
	Cd	114	139942.6	2.0	26.5678	1.569	5.9	ug/L	15	Standard
>	In	115	526017.9	3.9				ug/L	515400	Standard
	Sn	118	7243.4	3.7	1.0452	0.032	3.1	ug/L	1043	Standard
	Sb	123	117714.8	1.6	25.2129	1.285	5.1	ug/L	227	Standard
	Ba	135	327366.2	0.8	159.7393	7.528	4.7	ug/L	31	Standard
	Ce	140	1537481.7	1.2				ug/L	120	Standard
>	Tb	159	931910.0	4.2				ug/L	785817	Standard
	Ho	165	36582.4	1.2				ug/L	7	Standard
	Tl	203	213475.7	0.7	23.8226	1.044	4.4	ug/L	6	Standard
	Tl	205	175419.7	0.7	22.4638	0.769	3.4	ug/L	8	Standard
	Pb	206	334845.6	1.3	60.0496	2.994	5.0	ug/L	201	Standard
	Pb	207	283040.3	0.6	56.1331	2.483	4.4	ug/L	177	Standard
	Pb	208	1085063.5	0.5	53.5671	2.289	4.3	ug/L	722	Standard
	U	238	197136.0	1.2	25.6127	0.785	3.1	ug/L	8	Standard
>	Bi	209	498598.6	3.8				ug/L	444563	Standard

Sample ID: L1606145109S WG574663-04
Report Date/Time: Tuesday, July 05, 2016 10:27:21
Page 1

Approved: July 06, 2016

Na	23	3.3	86.6	**28.6197**	24.784	86.6	mg/L	0	Standard
Mg	24	646.7	16.9	**0.5579**	0.096	17.3	mg/L	38	Standard
K	39	175.0	20.0	**0.4284**	0.106	24.7	mg/L	33	Standard
Ca	43	210.0	17.2	**18.2405**	4.420	24.2	mg/L	48	Standard
Fe	54	20289.2	3.6	**11.8012**	0.292	2.5	mg/L	131	Standard
Fe	57	5335.9	2.2	**11.5671**	0.115	1.0	mg/L	153	Standard
Sc-1	45	36654.3	1.2				mg/L	26715	Standard
Cl	35	125111.1	0.6				ug/L	102686	Standard
Kr	83	2.7	43.3				ug/L	2	Standard
Br	81	6341.3	3.8				ug/L	1897	Standard
P	31	21325.0	1.2				ug/L	18491	Standard
S	34	5641.1	1.7				ug/L	3734	Standard
Sr	88	180.0	12.1				ug/L	47	Standard
C	12	1503.4	2.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	426.7	15.6				mg/L	23	Standard
Dy	164	55395.5	2.5				mg/L	23	Standard
Ho-1	165	36582.4	1.2				mg/L	7	Standard
Er	166	35110.6	0.4				mg/L	13	Standard
I	127	69658.4	5.3				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		154.263	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		109.150	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606145109S WG574663-04
Report Date/Time: Tuesday, July 05, 2016 10:27:21
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.060
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		112.155
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Cr 53 Upper, S, EEE	Cr	53	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606145109S WG574663-04
Report Date/Time: Tuesday, July 05, 2016 10:27:21
Page 3

Approved: July 06, 2016

Zn 66 Upper, S, EEE	Zn	66
Se-1 77 Upper, S, EEE	Se-1	77
Ba 135 Upper, S, EEE	Ba	135

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145110SD WG574663-05
Sample Date/Time: Tuesday, July 05, 2016 10:28:15
Number of Replicates: 3
Autosampler Position: 209
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92567.3	4.4				ug/L	59517	Standard
	Be	9	27448.2	1.3	**20.2975**	1.096	5.4	ug/L	10	Standard
	Al	27	3947408.5	1.4	**16.6926**	0.844	5.1	ug/L	1872	Standard
	Sc	45	37227.4	2.9				ug/L	26715	Standard
	Ti	47	22470.7	1.4	**85.5564**	0.807	0.9	ug/L	54	Standard
	V	51	787499.3	2.0	**98.9258**	2.683	2.7	ug/L	1419	Standard
	Cr	52	377991.5	1.4	**47.2948**	0.737	1.6	ug/L	7059	Standard
	Cr	53	186356.2	2.4	**189.9303**	6.848	3.6	ug/L	398	Standard
	Mn	55	2224101.6	2.0	**318.9399**	2.166	0.7	ug/L	824	Standard
	Co	59	213001.6	2.9	**28.4300**	0.204	0.7	ug/L	167	Standard
	Ni	60	102452.9	1.6	**49.5228**	0.586	1.2	ug/L	391	Standard
	Cu	65	168431.6	2.1	**81.3787**	0.641	0.8	ug/L	147	Standard
	Zn	66	130918.9	1.6	**107.5275**	0.961	0.9	ug/L	184	Standard
>	**Ge**	72	465960.3	2.2				ug/L	422998	Standard
	As	75	50506.2	2.6	**40.4629**	0.279	0.7	ug/L	-58	Standard
	Se	82	3972.9	1.3	**30.5627**	0.577	1.9	ug/L	15	Standard
	Se-1	77	15238.3	0.9	**181.8181**	4.788	2.6	ug/L	58	Standard
>	**Ga**	71	12636.9	5.1				mg/L	113	Standard
	Rb	85	34183.6	3.8				ug/L	27	Standard
	Y	89	934981.2	1.9				ug/L	383780	Standard
>	**Rh**	103	91.7	39.5				ug/L	18	Standard
	Mo	98	12089.5	2.1	**4.3684**	0.052	1.2	ug/L	31	Standard
	Ag	107	177713.5	0.9	**25.3798**	0.139	0.5	ug/L	84	Standard
	Cd	111	59831.1	3.0	**27.0406**	0.556	2.1	mg/L	4	Standard
	Cd	114	145986.8	3.5	**27.1872**	0.590	2.2	ug/L	15	Standard
>	**In**	115	535307.0	1.4				ug/L	515400	Standard
	Sn	118	7665.3	2.9	**1.0941**	0.052	4.8	ug/L	1043	Standard
	Sb	123	122141.1	0.8	**25.6760**	0.292	1.1	ug/L	227	Standard
	Ba	135	357831.6	1.4	**171.3677**	1.173	0.7	ug/L	31	Standard
	Ce	140	1664011.1	2.6				ug/L	120	Standard
>	**Tb**	159	942853.7	0.8				ug/L	785817	Standard
	Ho	165	39693.6	2.5				ug/L	7	Standard
	Tl	203	219374.2	1.1	**24.1748**	0.193	0.8	ug/L	6	Standard
	Tl	205	181310.6	2.9	**22.9312**	0.490	2.1	ug/L	8	Standard
	Pb	206	361102.5	1.7	**63.9442**	1.161	1.8	ug/L	201	Standard
	Pb	207	305063.7	1.7	**59.7456**	0.803	1.3	ug/L	177	Standard
	Pb	208	1165074.5	1.1	**56.8063**	0.948	1.7	ug/L	722	Standard
	U	238	205870.2	2.0	**26.4212**	0.390	1.5	ug/L	8	Standard
>	**Bi**	209	504359.6	1.0				ug/L	444563	Standard

Sample ID: L1606145110SD WG574663-05
Report Date/Time: Tuesday, July 05, 2016 10:30:32
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	641.7	11.4	**0.5457**	0.076	14.0	mg/L	38	Standard
K	39	143.3	14.1	**0.3220**	0.071	22.1	mg/L	33	Standard
Ca	43	233.3	16.2	**20.4787**	4.301	21.0	mg/L	48	Standard
Fe	54	23912.6	2.6	**13.7191**	0.336	2.5	mg/L	131	Standard
Fe	57	6194.6	4.8	**13.2925**	0.564	4.2	mg/L	153	Standard
> **Sc-1**	45	37227.4	2.9				mg/L	26715	Standard
Cl	35	124605.9	2.8				ug/L	102686	Standard
Kr	83	3.3	91.7				ug/L	2	Standard
Br	81	5787.8	8.8				ug/L	1897	Standard
P	31	22254.7	1.1				ug/L	18491	Standard
S	34	5691.1	4.1				ug/L	3734	Standard
Sr	88	145.0	18.2				ug/L	47	Standard
C	12	1630.1	3.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	570.0	14.4				mg/L	23	Standard
Dy	164	60514.1	3.8				mg/L	23	Standard
Ho-1	165	39693.6	2.5				mg/L	7	Standard
Er	166	38986.8	3.0				mg/L	13	Standard
I	127	74084.4	5.7				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		155.531	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.157	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145110SD WG574663-05
Report Date/Time: Tuesday, July 05, 2016 10:30:32
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.862
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.451
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Cr 53 Upper, S, EEE	Cr	53	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606145110SD WG574663-05
Report Date/Time: Tuesday, July 05, 2016 10:30:32
Page 3

Approved: July 06, 2016

Zn 66 Upper, S, EEE	Zn	66
Se-1 77 Upper, S, EEE	Se-1	77
Ba 135 Upper, S, EEE	Ba	135

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145101
Sample Date/Time: Tuesday, July 05, 2016 10:31:27
Number of Replicates: 3
Autosampler Position: 210
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96300.3	3.4				ug/L	59517	Standard
	Be	9	420.0	9.7	**0.2670**	0.019	7.0	ug/L	10	Standard
	Al	27	2367758.4	2.6	**9.6073**	0.083	0.9	ug/L	1872	Standard
	Sc	45	34126.7	0.7				ug/L	26715	Standard
	Ti	47	24224.4	0.8	**90.6734**	1.027	1.1	ug/L	54	Standard
	V	51	134625.6	3.4	**16.4696**	0.495	3.0	ug/L	1419	Standard
	Cr	52	100176.0	5.2	**11.5973**	0.644	5.5	ug/L	7059	Standard
	Cr	53	218029.7	14.7	**218.3883**	31.900	14.6	ug/L	398	Standard
	Mn	55	1246069.6	1.4	**175.6002**	3.333	1.9	ug/L	824	Standard
	Co	59	19635.0	1.5	**2.5539**	0.052	2.0	ug/L	167	Standard
	Ni	60	16067.5	1.5	**7.4648**	0.122	1.6	ug/L	391	Standard
	Cu	65	20297.6	1.6	**9.5712**	0.209	2.2	ug/L	147	Standard
	Zn	66	90111.9	1.7	**72.7009**	1.609	2.2	ug/L	184	Standard
>	**Ge**	72	474008.2	0.5				ug/L	422998	Standard
	As	75	10729.1	3.7	**8.4913**	0.334	3.9	ug/L	-58	Standard
	Se	82	74.1	24.9	**0.3575**	0.139	39.0	ug/L	15	Standard
	Se-1	77	15628.8	13.6	**183.2190**	24.701	13.5	ug/L	58	Standard
>	**Ga**	71	9844.9	2.3				mg/L	113	Standard
	Rb	85	75058.9	2.6				ug/L	27	Standard
	Y	89	496613.8	1.6				ug/L	383780	Standard
>	**Rh**	103	38.3	7.5				ug/L	18	Standard
	Mo	98	2422.5	2.0	**0.8423**	0.021	2.5	ug/L	31	Standard
	Ag	107	431.0	5.5	**0.0478**	0.004	7.4	ug/L	84	Standard
	Cd	111	826.3	3.8	**0.3584**	0.015	4.2	mg/L	4	Standard
	Cd	114	2066.0	2.3	**0.3689**	0.010	2.7	ug/L	15	Standard
>	**In**	115	551668.5	0.5				ug/L	515400	Standard
	Sn	118	6988.3	4.5	**0.9465**	0.045	4.8	ug/L	1043	Standard
	Sb	123	2722.2	8.4	**0.5392**	0.045	8.3	ug/L	227	Standard
	Ba	135	112861.5	0.6	**52.4311**	0.519	1.0	ug/L	31	Standard
	Ce	140	740222.4	0.5				ug/L	120	Standard
>	**Tb**	159	940061.7	0.9				ug/L	785817	Standard
	Ho	165	5439.3	2.8				ug/L	7	Standard
	Tl	203	855.0	6.5	**0.0902**	0.006	6.2	ug/L	6	Standard
	Tl	205	756.7	7.3	**0.0885**	0.006	7.2	ug/L	8	Standard
	Pb	206	282348.8	0.2	**49.7501**	0.303	0.6	ug/L	201	Standard
	Pb	207	237157.4	0.8	**46.2181**	0.627	1.4	ug/L	177	Standard
	Pb	208	908660.4	1.0	**44.0818**	0.650	1.5	ug/L	722	Standard
	U	238	4636.0	1.4	**0.5898**	0.010	1.7	ug/L	8	Standard
>	**Bi**	209	506760.7	0.6				ug/L	444563	Standard

Sample ID: L1606145101
Report Date/Time: Tuesday, July 05, 2016 10:33:44
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	138.3	17.1	**0.1024**	0.024	23.2	mg/L	38	Standard
K	39	191.7	24.1	**0.5278**	0.160	30.2	mg/L	33	Standard
Ca	43	85.0	15.6	**4.5864**	1.562	34.1	mg/L	48	Standard
Fe	54	11199.1	3.0	**6.9495**	0.163	2.4	mg/L	131	Standard
Fe	57	3090.3	4.2	**7.0129**	0.282	4.0	mg/L	153	Standard
Sc-1	45	34126.7	0.7				mg/L	26715	Standard
Cl	35	119518.9	0.7				ug/L	102686	Standard
Kr	83	2.7	86.6				ug/L	2	Standard
Br	81	2663.6	2.8				ug/L	1897	Standard
P	31	22528.4	3.2				ug/L	18491	Standard
S	34	6091.2	2.3				ug/L	3734	Standard
Sr	88	101.7	67.9				ug/L	47	Standard
C	12	383.3	9.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	170.0	10.2				mg/L	23	Standard
Dy	164	9346.6	4.0				mg/L	23	Standard
Ho-1	165	5439.3	2.8				mg/L	7	Standard
Er	166	5214.2	8.3				mg/L	13	Standard
I	127	32424.7	3.2				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		161.804	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		112.059	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145101
Report Date/Time: Tuesday, July 05, 2016 10:33:44
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		107.037
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.991
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Cr 53 Upper, S, EEE	Cr	53	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606145101
Report Date/Time: Tuesday, July 05, 2016 10:33:44
Page 3

Approved: July 06, 2016

Se-1 77 Upper, S, EEE Se-1 77

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145102
Sample Date/Time: Tuesday, July 05, 2016 10:34:38
Number of Replicates: 3
Autosampler Position: 211
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	95455.0	3.3				ug/L	59517	Standard
	Be	9	488.3	3.1	**0.3196**	0.023	7.1	ug/L	10	Standard
	Al	27	6729800.5	2.4	**27.5831**	0.993	3.6	ug/L	1872	Standard
	Sc	45	34343.9	3.1				ug/L	26715	Standard
	Ti	47	29782.6	2.0	**108.4368**	3.647	3.4	ug/L	54	Standard
	V	51	174788.9	3.0	**20.8308**	0.425	2.0	ug/L	1419	Standard
	Cr	52	146605.1	2.1	**16.9125**	0.463	2.7	ug/L	7059	Standard
	Cr	53	152022.1	9.9	**147.9448**	15.301	10.3	ug/L	398	Standard
	Mn	55	3975518.2	1.1	**544.8987**	11.406	2.1	ug/L	824	Standard
	Co	59	48242.3	0.2	**6.1346**	0.107	1.7	ug/L	167	Standard
	Ni	60	31076.2	1.0	**14.2120**	0.312	2.2	ug/L	391	Standard
	Cu	65	69088.5	0.2	**31.8512**	0.503	1.6	ug/L	147	Standard
	Zn	66	196946.8	0.4	**154.6413**	2.812	1.8	ug/L	184	Standard
>	**Ge**	72	487665.0	1.6				ug/L	422998	Standard
	As	75	10052.7	3.3	**7.7360**	0.126	1.6	ug/L	-58	Standard
	Se	82	100.5	8.6	**0.5373**	0.060	11.1	ug/L	15	Standard
	Se-1	77	12181.3	9.4	**138.6940**	13.419	9.7	ug/L	58	Standard
>	**Ga**	71	13002.2	3.0				mg/L	113	Standard
	Rb	85	139400.1	1.8				ug/L	27	Standard
	Y	89	543498.9	1.3				ug/L	383780	Standard
>	**Rh**	103	58.3	39.6				ug/L	18	Standard
	Mo	98	7908.6	2.5	**2.7773**	0.110	4.0	ug/L	31	Standard
	Ag	107	560.7	7.4	**0.0660**	0.006	9.8	ug/L	84	Standard
	Cd	111	885.6	4.5	**0.3854**	0.020	5.2	mg/L	4	Standard
	Cd	114	2255.4	2.1	**0.4043**	0.015	3.8	ug/L	15	Standard
>	**In**	115	550414.6	1.7				ug/L	515400	Standard
	Sn	118	10740.5	2.1	**1.5560**	0.019	1.2	ug/L	1043	Standard
	Sb	123	3508.6	7.0	**0.7010**	0.040	5.6	ug/L	227	Standard
	Ba	135	233884.5	1.0	**108.9317**	0.716	0.7	ug/L	31	Standard
	Ce	140	773942.7	0.2				ug/L	120	Standard
>	**Tb**	159	936947.5	1.0				ug/L	785817	Standard
	Ho	165	8690.8	1.9				ug/L	7	Standard
	Tl	203	1666.8	1.8	**0.1776**	0.007	3.7	ug/L	6	Standard
	Tl	205	1376.7	2.2	**0.1650**	0.005	3.0	ug/L	8	Standard
	Pb	206	174591.5	0.7	**30.4684**	0.696	2.3	ug/L	201	Standard
	Pb	207	143512.1	0.5	**27.6966**	0.652	2.4	ug/L	177	Standard
	Pb	208	564015.7	0.3	**27.0940**	0.434	1.6	ug/L	722	Standard
	U	238	10234.5	2.1	**1.2931**	0.032	2.5	ug/L	8	Standard
>	**Bi**	209	511498.1	1.9				ug/L	444563	Standard

Sample ID: L1606145102
Report Date/Time: Tuesday, July 05, 2016 10:36:55
Page 1

Approved: July 06, 2016

Na	23	3.3	86.6	**30.4304**	26.386	86.7	mg/L	0	Standard
Mg	24	231.7	12.6	**0.1933**	0.035	18.1	mg/L	38	Standard
K	39	413.3	26.3	**1.2849**	0.411	32.0	mg/L	33	Standard
Ca	43	188.3	14.6	**17.1374**	2.792	16.3	mg/L	48	Standard
Fe	54	21089.1	1.4	**13.1129**	0.341	2.6	mg/L	131	Standard
Fe	57	5754.4	1.9	**13.4013**	0.686	5.1	mg/L	153	Standard
Sc-1	45	34343.9	3.1				mg/L	26715	Standard
Cl	35	127275.9	2.5				ug/L	102686	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	3720.5	3.8				ug/L	1897	Standard
P	31	23304.6	0.3				ug/L	18491	Standard
S	34	6324.7	3.4				ug/L	3734	Standard
Sr	88	123.3	18.7				ug/L	47	Standard
C	12	446.7	26.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	176.7	14.2				mg/L	23	Standard
Dy	164	15071.2	2.7				mg/L	23	Standard
Ho-1	165	8690.8	1.9				mg/L	7	Standard
Er	166	8282.3	3.5				mg/L	13	Standard
I	127	47730.2	4.6				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		160.383	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.288	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145102
Report Date/Time: Tuesday, July 05, 2016 10:36:55
Page 2

Approved: July 06, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	106.794
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	115.056
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Ti 47 Upper, S, EEE	Ti	47	
Cr 53 Upper, S, EEE	Cr	53	

Sample ID: L1606145102
Report Date/Time: Tuesday, July 05, 2016 10:36:55
Page 3

Approved: July 06, 2016

Mn 55 Upper, S, EEE	Mn	55
Zn 66 Upper, S, EEE	Zn	66
Se-1 77 Upper, S, EEE	Se-1	77
Ba 135 Upper, S, EEE	Ba	135

Sample ID: L1606145102
Report Date/Time: Tuesday, July 05, 2016 10:36:55
Page 4

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145102PS WG575049-01
Sample Date/Time: Tuesday, July 05, 2016 10:37:50
Number of Replicates: 3
Autosampler Position: 212
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	92486.5	2.1				ug/L	59517	Standard
	Be	9	54039.6	2.5	**39.9902**	1.823	4.6	ug/L	10	Standard
	Al	27	6702719.9	3.6	**28.3315**	0.431	1.5	ug/L	1872	Standard
	Sc	45	33709.2	4.7				ug/L	26715	Standard
	Ti	47	29540.8	1.2	**111.4556**	1.217	1.1	ug/L	54	Standard
	V	51	606531.6	1.7	**75.3912**	0.314	0.4	ug/L	1419	Standard
	Cr	52	567176.4	1.3	**70.7450**	0.671	0.9	ug/L	7059	Standard
	Cr	53	188806.4	3.0	**190.4660**	3.113	1.6	ug/L	398	Standard
	Mn	55	4291846.8	1.5	**609.6425**	8.461	1.4	ug/L	824	Standard
	Co	59	434604.6	0.9	**57.4977**	1.668	2.9	ug/L	167	Standard
	Ni	60	139210.3	1.0	**66.7060**	0.799	1.2	ug/L	391	Standard
	Cu	65	178227.9	1.0	**85.2934**	1.566	1.8	ug/L	147	Standard
	Zn	66	258120.8	0.5	**210.1308**	4.737	2.3	ug/L	184	Standard
>	**Ge**	72	470547.5	2.2				ug/L	422998	Standard
	As	75	75543.5	1.9	**59.9211**	1.340	2.2	ug/L	-58	Standard
	Se	82	6892.2	1.4	**52.6663**	1.888	3.6	ug/L	15	Standard
	Se-1	77	15541.0	0.3	**183.6099**	3.476	1.9	ug/L	58	Standard
>	**Ga**	71	13265.8	1.7				mg/L	113	Standard
	Rb	85	137203.4	0.8				ug/L	27	Standard
	Y	89	529720.9	2.4				ug/L	383780	Standard
>	**Rh**	103	63.3	25.4				ug/L	18	Standard
	Mo	98	7951.5	2.9	**2.8798**	0.055	1.9	ug/L	31	Standard
	Ag	107	369070.7	0.4	**52.9279**	2.054	3.9	ug/L	84	Standard
	Cd	111	116570.1	1.0	**52.8942**	1.790	3.4	mg/L	4	Standard
	Cd	114	281010.1	0.2	**52.5513**	1.895	3.6	ug/L	15	Standard
>	**In**	115	533710.9	3.8				ug/L	515400	Standard
	Sn	118	10321.9	4.2	**1.5405**	0.022	1.5	ug/L	1043	Standard
	Sb	123	239957.0	1.1	**50.6654**	2.386	4.7	ug/L	227	Standard
	Ba	135	339166.4	0.2	**163.0629**	6.024	3.7	ug/L	31	Standard
	Ce	140	757093.0	0.6				ug/L	120	Standard
>	**Tb**	159	918140.9	2.3				ug/L	785817	Standard
	Ho	165	8654.1	3.7				ug/L	7	Standard
	Tl	203	466157.4	0.4	**51.0161**	1.276	2.5	ug/L	6	Standard
	Tl	205	384400.8	1.5	**48.2945**	1.556	3.2	ug/L	8	Standard
	Pb	206	473813.5	0.3	**83.3316**	2.208	2.6	ug/L	201	Standard
	Pb	207	406320.4	0.2	**79.0456**	2.340	3.0	ug/L	177	Standard
	Pb	208	1553678.1	1.6	**75.2238**	1.416	1.9	ug/L	722	Standard
	U	238	393828.7	1.0	**50.1983**	1.449	2.9	ug/L	8	Standard
>	**Bi**	209	508141.7	3.0				ug/L	444563	Standard

Sample ID: L1606145102PS WG575049-01
Report Date/Time: Tuesday, July 05, 2016 10:40:07
Page 1

Approved: July 06, 2016

Na	23	3.3	86.6	**31.0869**	27.005	86.9	mg/L	0	Standard
Mg	24	198.3	18.6	**0.1637**	0.035	21.4	mg/L	38	Standard
K	39	406.7	10.5	**1.2854**	0.179	13.9	mg/L	33	Standard
Ca	43	155.0	8.5	**13.4861**	1.915	14.2	mg/L	48	Standard
Fe	54	21393.9	2.9	**13.5718**	0.780	5.7	mg/L	131	Standard
Fe	57	5667.7	4.9	**13.4360**	0.276	2.1	mg/L	153	Standard
> **Sc-1**	45	33709.2	4.7				mg/L	26715	Standard
Cl	35	122143.7	1.1				ug/L	102686	Standard
Kr	83	2.7	108.3				ug/L	2	Standard
Br	81	3353.7	7.7				ug/L	1897	Standard
P	31	23266.2	1.4				ug/L	18491	Standard
S	34	5949.5	2.7				ug/L	3734	Standard
Sr	88	78.3	36.9				ug/L	47	Standard
C	12	416.7	6.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	156.7	3.7				mg/L	23	Standard
Dy	164	14716.3	3.4				mg/L	23	Standard
Ho-1	165	8654.1	3.7				mg/L	7	Standard
Er	166	7888.7	4.0				mg/L	13	Standard
I	127	47038.1	6.2				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		155.396	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		111.241	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145102PS WG575049-01
Report Date/Time: Tuesday, July 05, 2016 10:40:07
Page 2

Approved: July 06, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	103.553
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	114.301
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Ti 47 Upper, S, EEE	Ti	47	
Cr 53 Upper, S, EEE	Cr	53	

Sample ID: L1606145102PS WG575049-01
Report Date/Time: Tuesday, July 05, 2016 10:40:07
Page 3

Approved: July 06, 2016

Mn 55 Upper, S, EEE	Mn	55
Zn 66 Upper, S, EEE	Zn	66
Se-1 77 Upper, S, EEE	Se-1	77
Ba 135 Upper, S, EEE	Ba	135

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145102SDL WG575049-02
Sample Date/Time: Tuesday, July 05, 2016 10:41:01
Number of Replicates: 3
Autosampler Position: 213
Sample Description: 5
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	57964.0	0.6				ug/L	59517	Standard
	Be	9	111.7	2.6	**0.1008**	0.003	2.9	ug/L	10	Standard
	Al	27	1095254.7	1.1	**7.3789**	0.033	0.5	ug/L	1872	Standard
	Sc	45	28546.9	1.5				ug/L	26715	Standard
	Ti	47	5496.3	1.3	**22.0763**	0.684	3.1	ug/L	54	Standard
	V	51	31405.8	2.5	**4.0242**	0.171	4.3	ug/L	1419	Standard
	Cr	52	30164.7	1.1	**3.1202**	0.057	1.8	ug/L	7059	Standard
	Cr	53	40515.8	1.3	**43.5878**	1.037	2.4	ug/L	398	Standard
	Mn	55	729144.7	0.3	**111.1254**	2.253	2.0	ug/L	824	Standard
	Co	59	8726.8	1.1	**1.2153**	0.041	3.4	ug/L	167	Standard
	Ni	60	6092.9	0.9	**2.9447**	0.045	1.5	ug/L	391	Standard
	Cu	65	13417.3	0.9	**6.8221**	0.107	1.6	ug/L	147	Standard
	Zn	66	41065.9	1.2	**35.7658**	0.867	2.4	ug/L	184	Standard
>	Ge	72	438215.2	2.3				ug/L	422998	Standard
	As	75	1917.2	6.6	**1.6850**	0.145	8.6	ug/L	-58	Standard
	Se	82	32.6	6.4	**0.0620**	0.014	22.6	ug/L	15	Standard
	Se-1	77	2564.6	2.2	**31.9793**	0.528	1.7	ug/L	58	Standard
>	Ga	71	2510.2	5.3				mg/L	113	Standard
	Rb	85	24840.3	1.6				ug/L	27	Standard
	Y	89	392893.1	1.7				ug/L	383780	Standard
>	Rh	103	16.7	75.5				ug/L	18	Standard
	Mo	98	1401.8	4.4	**0.5331**	0.032	5.9	ug/L	31	Standard
	Ag	107	213.7	10.6	**0.0206**	0.003	15.1	ug/L	84	Standard
	Cd	111	183.9	9.9	**0.0845**	0.008	9.5	mg/L	4	Standard
	Cd	114	476.8	11.7	**0.0901**	0.010	10.7	ug/L	15	Standard
>	In	115	501582.7	1.5				ug/L	515400	Standard
	Sn	118	2480.2	11.4	**0.2597**	0.056	21.6	ug/L	1043	Standard
	Sb	123	2141.9	24.9	**0.4649**	0.123	26.4	ug/L	227	Standard
	Ba	135	42972.9	0.5	**21.9469**	0.425	1.9	ug/L	31	Standard
	Ce	140	142041.1	0.9				ug/L	120	Standard
>	Tb	159	806653.6	1.4				ug/L	785817	Standard
	Ho	165	1461.7	2.6				ug/L	7	Standard
	Tl	203	382.7	6.2	**0.0423**	0.003	6.3	ug/L	6	Standard
	Tl	205	305.0	24.1	**0.0352**	0.010	27.4	ug/L	8	Standard
	Pb	206	31166.7	1.0	**5.9721**	0.086	1.4	ug/L	201	Standard
	Pb	207	25233.3	0.5	**5.3414**	0.081	1.5	ug/L	177	Standard
	Pb	208	105054.1	1.2	**5.5370**	0.086	1.5	ug/L	722	Standard
	U	238	1882.1	5.1	**0.2607**	0.013	4.9	ug/L	8	Standard
>	Bi	209	462925.1	1.5				ug/L	444563	Standard

Sample ID: L1606145102SDL WG575049-02
Report Date/Time: Tuesday, July 05, 2016 10:43:18
Page 1

Approved: July 06, 2016

Na	23	3.3	86.6	**36.5066**	31.621	86.6	mg/L	0	Standard
Mg	24	40.0	45.1	**0.0136**	0.022	160.1	mg/L	38	Standard
K	39	98.3	19.3	**0.2737**	0.078	28.6	mg/L	33	Standard
Ca	43	73.3	10.4	**4.9246**	1.170	23.8	mg/L	48	Standard
Fe	54	3790.7	3.4	**2.7433**	0.104	3.8	mg/L	131	Standard
Fe	57	1205.0	3.0	**3.0135**	0.147	4.9	mg/L	153	Standard
> **Sc-1**	45	28546.9	1.5				mg/L	26715	Standard
Cl	35	102927.1	0.4				ug/L	102686	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	1773.4	4.3				ug/L	1897	Standard
P	31	8959.3	0.9				ug/L	18491	Standard
S	34	4902.5	3.1				ug/L	3734	Standard
Sr	88	51.7	24.4				ug/L	47	Standard
C	12	66.7	31.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	40.0	25.0				mg/L	23	Standard
Dy	164	2723.5	4.1				mg/L	23	Standard
Ho-1	165	1461.7	2.6				mg/L	7	Standard
Er	166	1393.4	9.5				mg/L	13	Standard
I	127	36296.8	0.9				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		97.391	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.598	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145102SDL WG575049-02
Report Date/Time: Tuesday, July 05, 2016 10:43:18
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	97.319	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	104.130	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606145102SDL WG575049-02
Report Date/Time: Tuesday, July 05, 2016 10:43:18
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145102SDL WG575049-02
Sample Date/Time: Tuesday, July 05, 2016 10:44:13
Number of Replicates: 3
Autosampler Position: 214
Sample Description: 25
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	49243.3	3.3				ug/L	59517	Standard
	Be	9	26.7	39.0	**0.0058**	0.014	237.4	ug/L	10	Standard
	Al	27	212721.3	1.3	**1.6783**	0.042	2.5	ug/L	1872	Standard
	Sc	45	27433.2	2.0				ug/L	26715	Standard
	Ti	47	1107.7	3.1	**4.2771**	0.178	4.2	ug/L	54	Standard
	V	51	7348.6	1.8	**0.8095**	0.030	3.7	ug/L	1419	Standard
	Cr	52	10131.1	1.7	**0.4074**	0.025	6.1	ug/L	7059	Standard
	Cr	53	15222.6	3.6	**16.2080**	0.438	2.7	ug/L	398	Standard
	Mn	55	145385.2	0.9	**22.1749**	0.340	1.5	ug/L	824	Standard
	Co	59	1900.8	1.5	**0.2465**	0.003	1.3	ug/L	167	Standard
	Ni	60	1466.7	0.8	**0.5613**	0.007	1.2	ug/L	391	Standard
	Cu	65	2781.6	1.2	**1.3599**	0.036	2.7	ug/L	147	Standard
	Zn	66	9116.1	1.1	**7.8588**	0.045	0.6	ug/L	184	Standard
>	Ge	72	435735.1	1.6				ug/L	422998	Standard
	As	75	307.5	38.6	**0.3158**	0.105	33.4	ug/L	-58	Standard
	Se	82	18.6	74.4	**-0.0509**	0.117	229.4	ug/L	15	Standard
	Se-1	77	847.4	3.1	**10.1822**	0.435	4.3	ug/L	58	Standard
>	Ga	71	563.3	8.2				mg/L	113	Standard
	Rb	85	5184.2	0.5				ug/L	27	Standard
	Y	89	384411.6	4.5				ug/L	383780	Standard
>	Rh	103	16.7	45.8				ug/L	18	Standard
	Mo	98	334.4	9.1	**0.1194**	0.009	7.2	ug/L	31	Standard
	Ag	107	118.0	13.2	**0.0059**	0.003	44.4	ug/L	84	Standard
	Cd	111	48.7	10.3	**0.0193**	0.003	15.4	mg/L	4	Standard
	Cd	114	106.3	20.1	**0.0164**	0.004	25.0	ug/L	15	Standard
>	In	115	504541.4	2.8				ug/L	515400	Standard
	Sn	118	885.0	33.3	**-0.0256**	0.047	184.8	ug/L	1043	Standard
	Sb	123	1059.7	23.7	**0.2192**	0.050	22.7	ug/L	227	Standard
	Ba	135	8479.0	0.7	**4.2871**	0.116	2.7	ug/L	31	Standard
	Ce	140	27329.6	0.4				ug/L	120	Standard
>	Tb	159	780954.8	2.4				ug/L	785817	Standard
	Ho	165	301.7	20.5				ug/L	7	Standard
	Tl	203	128.7	11.8	**0.0125**	0.002	16.3	ug/L	6	Standard
	Tl	205	91.7	30.0	**0.0064**	0.004	63.8	ug/L	8	Standard
	Pb	206	6174.9	1.4	**1.1929**	0.007	0.6	ug/L	201	Standard
	Pb	207	5019.5	1.1	**1.0657**	0.013	1.2	ug/L	177	Standard
	Pb	208	21403.9	1.3	**1.1335**	0.003	0.3	ug/L	722	Standard
	U	238	365.7	5.3	**0.0506**	0.003	5.0	ug/L	8	Standard
>	Bi	209	445597.3	1.1				ug/L	444563	Standard

Sample ID: L1606145102SDL WG575049-02
Report Date/Time: Tuesday, July 05, 2016 10:46:30
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	31.7	18.2	**0.0051**	0.007	142.5	mg/L	38	Standard
K	39	48.3	15.8	**0.0770**	0.034	44.2	mg/L	33	Standard
Ca	43	46.7	34.4	**1.2487**	2.341	187.5	mg/L	48	Standard
Fe	54	833.8	6.6	**0.5377**	0.035	6.5	mg/L	131	Standard
Fe	57	361.7	13.3	**0.6103**	0.152	25.0	mg/L	153	Standard
> **Sc-1**	45	27433.2	2.0				mg/L	26715	Standard
Cl	35	99619.1	1.7				ug/L	102686	Standard
Kr	83	3.0	115.5				ug/L	2	Standard
Br	81	1573.4	5.8				ug/L	1897	Standard
P	31	6636.5	4.3				ug/L	18491	Standard
S	34	4434.0	1.9				ug/L	3734	Standard
Sr	88	80.0	16.5				ug/L	47	Standard
C	12	73.3	20.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	114.6				mg/L	23	Standard
Dy	164	495.0	12.2				mg/L	23	Standard
Ho-1	165	301.7	20.5				mg/L	7	Standard
Er	166	313.3	4.9				mg/L	13	Standard
I	127	34170.2	2.9				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		82.739	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		103.011	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145102SDL WG575049-02
Report Date/Time: Tuesday, July 05, 2016 10:46:30
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	97.893	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	100.233	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: L1606145102SDL WG575049-02
Report Date/Time: Tuesday, July 05, 2016 10:46:30
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, July 05, 2016 10:47:26
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	51443.9	2.3				ug/L	59517	Standard
	Be	9	39172.3	3.7	52.1373	3.047	5.8	ug/L	10	Standard
	Al	27	7091306.9	1.3	53.9261	1.288	2.4	ug/L	1872	Standard
	Sc	45	29433.6	6.4				ug/L	26715	Standard
	Ti	47	25883.1	1.7	99.6649	1.934	1.9	ug/L	54	Standard
	V	51	378476.2	1.5	47.9762	1.568	3.3	ug/L	1419	Standard
	Cr	52	388291.0	1.7	49.1588	1.543	3.1	ug/L	7059	Standard
	Cr	53	51136.2	2.2	52.3946	2.154	4.1	ug/L	398	Standard
	Mn	55	339015.4	1.1	49.0540	1.615	3.3	ug/L	824	Standard
	Co	59	351131.2	0.8	47.4186	1.464	3.1	ug/L	167	Standard
	Ni	60	101735.2	1.9	49.7282	1.809	3.6	ug/L	391	Standard
	Cu	65	100071.5	2.1	48.8595	1.575	3.2	ug/L	147	Standard
	Zn	66	58828.3	1.8	48.7660	1.404	2.9	ug/L	184	Standard
>	Ge	72	461081.6	3.6				ug/L	422998	Standard
	As	75	58960.5	1.4	47.7594	1.382	2.9	ug/L	-58	Standard
	Se	82	5982.2	1.4	46.6329	1.288	2.8	ug/L	15	Standard
	Se-1	77	4323.3	1.9	51.6659	1.569	3.0	ug/L	58	Standard
>	Ga	71	116.7	13.1				mg/L	113	Standard
	Rb	85	950.0	7.1				ug/L	27	Standard
	Y	89	397802.5	3.0				ug/L	383780	Standard
>	Rh	103	41.7	25.0				ug/L	18	Standard
	Mo	98	260548.4	1.4	95.0668	1.820	1.9	ug/L	31	Standard
	Ag	107	334433.8	1.0	48.1515	1.472	3.1	ug/L	84	Standard
	Cd	111	106832.0	0.8	48.6756	1.420	2.9	mg/L	4	Standard
	Cd	114	266119.3	1.2	49.9600	1.063	2.1	ug/L	15	Standard
>	In	115	531351.4	2.8				ug/L	515400	Standard
	Sn	118	290753.5	1.2	48.5577	1.676	3.5	ug/L	1043	Standard
	Sb	123	235402.2	0.6	49.8905	1.407	2.8	ug/L	227	Standard
	Ba	135	99526.5	0.7	48.0193	1.003	2.1	ug/L	31	Standard
	Ce	140	1210.1	40.8				ug/L	120	Standard
>	Tb	159	803321.7	2.7				ug/L	785817	Standard
	Ho	165	10.0	50.0				ug/L	7	Standard
	Tl	203	401152.8	0.4	48.9599	1.820	3.7	ug/L	6	Standard
	Tl	205	347377.0	0.4	48.6673	1.791	3.7	ug/L	8	Standard
	Pb	206	247960.2	0.4	48.6117	1.723	3.5	ug/L	201	Standard
	Pb	207	226113.7	0.4	49.0328	1.786	3.6	ug/L	177	Standard
	Pb	208	918814.9	0.9	49.6084	2.095	4.2	ug/L	722	Standard
	U	238	341289.7	0.5	48.5107	1.791	3.7	ug/L	8	Standard
>	Bi	209	455797.4	3.3				ug/L	444563	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 10:49:43
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3965.5	1.6	**4.4967**	0.254	5.6	mg/L	38	Standard
K	39	1118.4	8.8	**4.3111**	0.182	4.2	mg/L	33	Standard
Ca	43	58.3	48.7	**2.3697**	3.619	152.7	mg/L	48	Standard
Fe	54	6650.7	1.6	**4.7651**	0.377	7.9	mg/L	131	Standard
Fe	57	1871.8	6.0	**4.7828**	0.034	0.7	mg/L	153	Standard
Sc-1	45	29433.6	6.4				mg/L	26715	Standard
Cl	35	115258.8	1.3				ug/L	102686	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	1830.1	3.9				ug/L	1897	Standard
P	31	20310.3	0.3				ug/L	18491	Standard
S	34	5012.5	2.8				ug/L	3734	Standard
Sr	88	66.7	15.6				ug/L	47	Standard
C	12	40.0	43.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	23	Standard
Dy	164	19.2	47.1				mg/L	23	Standard
Ho-1	165	10.0	50.0				mg/L	7	Standard
Er	166	16.7	124.9				mg/L	13	Standard
I	127	17883.9	1.0				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	104.275		
Al	27	107.852		
Sc	45			
Ti	47	99.665		
V	51	95.952		
Cr	52	98.318		
Cr	53			
Mn	55	98.108		
Co	59	94.837		
Ni	60	99.456		
Cu	65	97.719		
Zn	66	97.532		
Ge	72		109.003	
As	75	95.519		
Se	82	93.266		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 10:49:43
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	95.067		
Ag	107	96.303		
Cd	111	97.351		
Cd	114			
> In	115			103.095
Sn	118	97.115		
Sb	123	99.781		
Ba	135	96.039		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	97.920		
Tl	205			
Pb	206			
Pb	207			
Pb	208	99.217		
U	238	97.021		
> Bi	209			102.527
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 10:49:43
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, July 05, 2016 10:50:37
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	52763.4	0.9				ug/L	59517	Standard
	Be	9	23.3	32.7	-0.0008	0.010	1314.7	ug/L	10	Standard
	Al	27	2333.5	18.2	0.0049	0.003	66.6	ug/L	1872	Standard
	Sc	45	28528.5	2.4				ug/L	26715	Standard
	Ti	47	36.7	23.8	-0.1009	0.036	35.4	ug/L	54	Standard
	V	51	1557.6	15.5	0.0243	0.029	119.3	ug/L	1419	Standard
	Cr	52	6897.2	1.6	-0.0594	0.009	15.4	ug/L	7059	Standard
	Cr	53	3362.0	2.2	3.1623	0.052	1.6	ug/L	398	Standard
	Mn	55	915.7	8.9	0.0057	0.010	183.1	ug/L	824	Standard
	Co	59	250.7	19.1	0.0098	0.006	65.3	ug/L	167	Standard
	Ni	60	310.0	6.8	-0.0432	0.013	29.5	ug/L	391	Standard
	Cu	65	150.0	9.3	-0.0036	0.006	170.0	ug/L	147	Standard
	Zn	66	266.0	3.4	0.0684	0.006	9.3	ug/L	184	Standard
>	Ge	72	447686.1	1.3				ug/L	422998	Standard
	As	75	-73.6	46.7	-0.0095	0.028	297.5	ug/L	-58	Standard
	Se	82	14.3	31.0	-0.0912	0.035	38.2	ug/L	15	Standard
	Se-1	77	251.3	11.2	2.4627	0.338	13.7	ug/L	58	Standard
>	Ga	71	81.7	9.4				mg/L	113	Standard
	Rb	85	30.0	33.3				ug/L	27	Standard
	Y	89	389776.6	0.4				ug/L	383780	Standard
>	Rh	103	15.0	33.3				ug/L	18	Standard
	Mo	98	215.1	11.9	0.0702	0.008	11.1	ug/L	31	Standard
	Ag	107	131.7	9.6	0.0071	0.002	22.0	ug/L	84	Standard
	Cd	111	19.9	27.5	0.0050	0.002	46.8	mg/L	4	Standard
	Cd	114	56.2	23.6	0.0061	0.002	38.5	ug/L	15	Standard
>	In	115	526258.8	2.4				ug/L	515400	Standard
	Sn	118	1383.4	28.2	0.0542	0.072	133.2	ug/L	1043	Standard
	Sb	123	1040.9	27.2	0.2074	0.065	31.6	ug/L	227	Standard
	Ba	135	56.7	43.3	0.0037	0.011	305.2	ug/L	31	Standard
	Ce	140	378.3	12.3				ug/L	120	Standard
>	Tb	159	796142.1	1.4				ug/L	785817	Standard
	Ho	165	10.0	50.0				ug/L	7	Standard
	Tl	203	119.0	10.2	0.0108	0.001	13.7	ug/L	6	Standard
	Tl	205	113.3	12.7	0.0090	0.002	20.6	ug/L	8	Standard
	Pb	206	279.3	17.7	0.0088	0.009	102.9	ug/L	201	Standard
	Pb	207	222.7	25.2	0.0001	0.012	15025.6	ug/L	177	Standard
	Pb	208	993.7	11.3	0.0047	0.006	116.9	ug/L	722	Standard
	U	238	65.3	18.5	0.0067	0.002	24.2	ug/L	8	Standard
>	Bi	209	458480.9	1.1				ug/L	444563	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 10:52:54
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**17.8060**	30.832	173.2	mg/L	0	Standard
Mg	24	28.3	10.2	**-0.0003**	0.003	905.5	mg/L	38	Standard
K	39	18.3	15.7	**-0.0541**	0.011	21.2	mg/L	33	Standard
Ca	43	31.7	9.1	**-1.2176**	0.312	25.6	mg/L	48	Standard
Fe	54	164.2	17.0	**0.0077**	0.023	303.7	mg/L	131	Standard
Fe	57	215.0	6.2	**0.1423**	0.047	33.2	mg/L	153	Standard
> **Sc-1**	45	28528.5	2.4				mg/L	26715	Standard
Cl	35	113758.1	0.6				ug/L	102686	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	1780.1	3.7				ug/L	1897	Standard
P	31	19746.2	1.4				ug/L	18491	Standard
S	34	4789.1	3.4				ug/L	3734	Standard
Sr	88	75.0	23.1				ug/L	47	Standard
C	12	56.7	44.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	23	Standard
Dy	164	9.0	9.2				mg/L	23	Standard
Ho-1	165	10.0	50.0				mg/L	7	Standard
Er	166	20.0	86.6				mg/L	13	Standard
I	127	11247.5	2.2				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		105.837	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 10:52:54
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.107
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.131
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
QC Std 7	Sb	123	

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 10:52:54
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145103
Sample Date/Time: Tuesday, July 05, 2016 10:53:50
Number of Replicates: 3
Autosampler Position: 215
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	96135.7	0.5				ug/L	59517	Standard
	Be	9	441.7	15.3	**0.2835**	0.050	17.5	ug/L	10	Standard
	Al	27	1899414.0	2.6	**7.7171**	0.235	3.0	ug/L	1872	Standard
	Sc	45	31502.7	2.3				ug/L	26715	Standard
	Ti	47	25531.2	2.1	**98.1864**	1.662	1.7	ug/L	54	Standard
	V	51	134345.2	6.2	**16.8817**	0.847	5.0	ug/L	1419	Standard
	Cr	52	120049.7	5.6	**14.4950**	0.617	4.3	ug/L	7059	Standard
	Cr	53	180887.5	11.3	**185.8515**	17.788	9.6	ug/L	398	Standard
	Mn	55	1694571.4	1.2	**245.3954**	5.740	2.3	ug/L	824	Standard
	Co	59	21225.2	1.9	**2.8384**	0.001	0.0	ug/L	167	Standard
	Ni	60	17246.1	2.6	**8.2498**	0.065	0.8	ug/L	391	Standard
	Cu	65	22767.1	1.8	**11.0388**	0.017	0.2	ug/L	147	Standard
	Zn	66	119572.4	2.0	**99.1503**	1.607	1.6	ug/L	184	Standard
>	**Ge**	72	461460.7	1.9				ug/L	422998	Standard
	As	75	5444.8	8.1	**4.4484**	0.294	6.6	ug/L	-58	Standard
	Se	82	77.6	3.9	**0.4009**	0.019	4.7	ug/L	15	Standard
	Se-1	77	15105.3	13.8	**181.6526**	21.923	12.1	ug/L	58	Standard
>	**Ga**	71	10466.9	3.1				mg/L	113	Standard
	Rb	85	79449.5	2.0				ug/L	27	Standard
	Y	89	453913.1	1.4				ug/L	383780	Standard
>	**Rh**	103	45.0	48.4				ug/L	18	Standard
	Mo	98	4163.6	3.0	**1.5365**	0.022	1.4	ug/L	31	Standard
	Ag	107	565.7	36.6	**0.0706**	0.029	40.7	ug/L	84	Standard
	Cd	111	905.2	6.7	**0.4152**	0.022	5.2	mg/L	4	Standard
	Cd	114	2285.7	5.6	**0.4318**	0.017	4.0	ug/L	15	Standard
>	**In**	115	522134.8	1.7				ug/L	515400	Standard
	Sn	118	11225.8	3.7	**1.7327**	0.054	3.1	ug/L	1043	Standard
	Sb	123	3009.2	12.2	**0.6318**	0.069	10.9	ug/L	227	Standard
	Ba	135	100468.8	2.1	**49.3104**	0.570	1.2	ug/L	31	Standard
	Ce	140	633585.7	1.9				ug/L	120	Standard
>	**Tb**	159	901421.0	1.9				ug/L	785817	Standard
	Ho	165	4162.2	5.6				ug/L	7	Standard
	Tl	203	1260.4	18.2	**0.1368**	0.025	18.0	ug/L	6	Standard
	Tl	205	1115.0	20.8	**0.1358**	0.029	21.0	ug/L	8	Standard
	Pb	206	328394.9	1.1	**58.8063**	0.550	0.9	ug/L	201	Standard
	Pb	207	271769.4	2.1	**53.8221**	0.890	1.7	ug/L	177	Standard
	Pb	208	1036625.3	1.0	**51.1073**	0.173	0.3	ug/L	722	Standard
	U	238	4954.8	7.4	**0.6405**	0.043	6.8	ug/L	8	Standard
>	**Bi**	209	498699.5	0.7				ug/L	444563	Standard

Sample ID: L1606145103
Report Date/Time: Tuesday, July 05, 2016 10:56:07
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	125.0	10.6	**0.0998**	0.017	17.3	mg/L	38	Standard
K	39	251.7	8.0	**0.8046**	0.062	7.7	mg/L	33	Standard
Ca	43	71.7	26.4	**3.6957**	2.561	69.3	mg/L	48	Standard
Fe	54	13368.3	3.5	**9.0209**	0.197	2.2	mg/L	131	Standard
Fe	57	3707.1	6.7	**9.2583**	0.592	6.4	mg/L	153	Standard
> **Sc-1**	45	31502.7	2.3				mg/L	26715	Standard
Cl	35	112834.0	1.9				ug/L	102686	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	2756.9	3.6				ug/L	1897	Standard
P	31	21984.3	2.1				ug/L	18491	Standard
S	34	5467.7	1.7				ug/L	3734	Standard
Sr	88	96.7	3.0				ug/L	47	Standard
C	12	416.7	23.1				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	136.7	27.7				mg/L	23	Standard
Dy	164	7702.9	4.4				mg/L	23	Standard
Ho-1	165	4162.2	5.6				mg/L	7	Standard
Er	166	4157.2	5.4				mg/L	13	Standard
I	127	34517.7	6.7				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		161.527	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		109.093	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145103
Report Date/Time: Tuesday, July 05, 2016 10:56:07
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.307
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		112.178
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Cr 53 Upper, S, EEE	Cr	53	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606145103
Report Date/Time: Tuesday, July 05, 2016 10:56:07
Page 3

Approved: July 06, 2016

Se-1 77 Upper, S, EEE Se-1 77

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145104
Sample Date/Time: Tuesday, July 05, 2016 10:57:02
Number of Replicates: 3
Autosampler Position: 216
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	91292.6	2.2				ug/L	59517	Standard
	Be	9	395.0	1.3	**0.2652**	0.010	3.9	ug/L	10	Standard
	Al	27	9819055.1	1.7	**42.0648**	0.383	0.9	ug/L	1872	Standard
	Sc	45	32252.6	1.6				ug/L	26715	Standard
	Ti	47	24196.3	0.3	**91.9481**	1.097	1.2	ug/L	54	Standard
	V	51	123723.8	2.5	**15.3573**	0.518	3.4	ug/L	1419	Standard
	Cr	52	104402.9	2.1	**12.3305**	0.473	3.8	ug/L	7059	Standard
	Cr	53	109957.5	19.5	**111.8418**	23.600	21.1	ug/L	398	Standard
	Mn	55	1745235.6	1.8	**249.6939**	1.366	0.5	ug/L	824	Standard
	Co	59	28619.0	1.2	**3.7905**	0.016	0.4	ug/L	167	Standard
	Ni	60	20456.1	0.9	**9.7078**	0.231	2.4	ug/L	391	Standard
	Cu	65	24146.9	2.2	**11.5780**	0.428	3.7	ug/L	147	Standard
	Zn	66	125347.7	0.8	**102.7332**	2.027	2.0	ug/L	184	Standard
>	**Ge**	72	466950.8	1.4				ug/L	422998	Standard
	As	75	5697.4	7.9	**4.5993**	0.326	7.1	ug/L	-58	Standard
	Se	82	78.7	8.4	**0.4018**	0.048	11.8	ug/L	15	Standard
	Se-1	77	9103.1	13.7	**108.2312**	16.582	15.3	ug/L	58	Standard
>	**Ga**	71	9131.1	1.4				mg/L	113	Standard
	Rb	85	76169.3	1.3				ug/L	27	Standard
	Y	89	475112.5	1.9				ug/L	383780	Standard
>	**Rh**	103	20.0	25.0				ug/L	18	Standard
	Mo	98	5031.6	1.4	**1.8226**	0.023	1.3	ug/L	31	Standard
	Ag	107	458.7	20.3	**0.0539**	0.013	25.0	ug/L	84	Standard
	Cd	111	769.3	2.5	**0.3456**	0.014	4.0	mg/L	4	Standard
	Cd	114	1978.5	10.3	**0.3661**	0.041	11.1	ug/L	15	Standard
>	**In**	115	532524.4	1.6				ug/L	515400	Standard
	Sn	118	8263.9	7.0	**1.1996**	0.075	6.2	ug/L	1043	Standard
	Sb	123	2736.8	8.9	**0.5620**	0.046	8.1	ug/L	227	Standard
	Ba	135	96479.9	0.9	**46.4314**	0.314	0.7	ug/L	31	Standard
	Ce	140	630403.5	2.4				ug/L	120	Standard
>	**Tb**	159	908285.0	0.8				ug/L	785817	Standard
	Ho	165	5554.4	1.9				ug/L	7	Standard
	Tl	203	1548.1	7.1	**0.1672**	0.009	5.6	ug/L	6	Standard
	Tl	205	1266.7	3.2	**0.1537**	0.003	1.9	ug/L	8	Standard
	Pb	206	242583.9	0.2	**43.0094**	0.811	1.9	ug/L	201	Standard
	Pb	207	201849.3	0.6	**39.5807**	0.878	2.2	ug/L	177	Standard
	Pb	208	795835.1	0.9	**38.8431**	0.400	1.0	ug/L	722	Standard
	U	238	5579.0	1.8	**0.7146**	0.002	0.3	ug/L	8	Standard
>	**Bi**	209	503660.9	1.9				ug/L	444563	Standard

Sample ID: L1606145104
Report Date/Time: Tuesday, July 05, 2016 10:59:18
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	111.7	18.1	**0.0826**	0.022	26.5	mg/L	38	Standard
K	39	205.0	10.6	**0.6134**	0.067	10.9	mg/L	33	Standard
Ca	43	158.3	25.3	**14.7113**	4.889	33.2	mg/L	48	Standard
Fe	54	14301.5	1.7	**9.4348**	0.276	2.9	mg/L	131	Standard
Fe	57	4067.2	3.4	**9.9615**	0.502	5.0	mg/L	153	Standard
> **Sc-1**	45	32252.6	1.6				mg/L	26715	Standard
Cl	35	125056.7	0.7				ug/L	102686	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	2540.2	11.1				ug/L	1897	Standard
P	31	21291.6	1.9				ug/L	18491	Standard
S	34	5712.7	2.7				ug/L	3734	Standard
Sr	88	96.7	7.9				ug/L	47	Standard
C	12	350.0	10.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	123.3	54.0				mg/L	23	Standard
Dy	164	9524.0	2.9				mg/L	23	Standard
Ho-1	165	5554.4	1.9				mg/L	7	Standard
Er	166	5130.9	3.9				mg/L	13	Standard
I	127	29181.4	3.9				mg/L	12398	Standard

QC Calculated Values

	Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
>	Li	6		153.390	
	Be	9			
	Al	27			
	Sc	45			
	Ti	47			
	V	51			
	Cr	52			
	Cr	53			
	Mn	55			
	Co	59			
	Ni	60			
	Cu	65			
	Zn	66			
>	Ge	72		110.391	
	As	75			
	Se	82			
	Se-1	77			
>	Ga	71			

Sample ID: L1606145104
Report Date/Time: Tuesday, July 05, 2016 10:59:18
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.323
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.294
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Cr 53 Upper, S, EEE	Cr	53	
Mn 55 Upper, S, EEE	Mn	55	

Sample ID: L1606145104
Report Date/Time: Tuesday, July 05, 2016 10:59:18
Page 3

Approved: July 06, 2016

Zn 66 Upper, S, EEE	Zn	66
Se-1 77 Upper, S, EEE	Se-1	77

Sample ID: L1606145104
Report Date/Time: Tuesday, July 05, 2016 10:59:18
Page 4

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145105
Sample Date/Time: Tuesday, July 05, 2016 11:00:13
Number of Replicates: 3
Autosampler Position: 217
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	108631.5	1.8				ug/L	59517	Standard
	Be	9	5212.6	1.8	**3.2539**	0.069	2.1	ug/L	10	Standard
	Al	27	3122829.6	1.3	**11.2336**	0.101	0.9	ug/L	1872	Standard
	Sc	45	39190.6	1.9				ug/L	26715	Standard
	Ti	47	137469.5	1.0	**500.3092**	6.619	1.3	ug/L	54	Standard
	V	51	366725.2	5.0	**43.8169**	2.017	4.6	ug/L	1419	Standard
	Cr	52	689707.4	1.9	**83.0086**	1.642	2.0	ug/L	7059	Standard
	Cr	53	253610.4	10.0	**246.6117**	25.946	10.5	ug/L	398	Standard
	Mn	55	865524.6	0.8	**118.2866**	1.664	1.4	ug/L	824	Standard
	Co	59	64362.7	0.2	**8.1759**	0.061	0.7	ug/L	167	Standard
	Ni	60	60072.2	0.5	**27.6005**	0.122	0.4	ug/L	391	Standard
	Cu	65	69465.5	0.6	**31.9592**	0.188	0.6	ug/L	147	Standard
	Zn	66	222603.4	0.9	**174.4424**	1.635	0.9	ug/L	184	Standard
>	**Ge**	72	488604.3	0.9				ug/L	422998	Standard
	As	75	19640.7	4.5	**15.0437**	0.806	5.4	ug/L	-58	Standard
	Se	82	157.4	9.5	**0.9557**	0.102	10.6	ug/L	15	Standard
	Se-1	77	19281.7	11.6	**219.5886**	27.202	12.4	ug/L	58	Standard
>	**Ga**	71	30325.3	1.1				mg/L	113	Standard
	Rb	85	276639.3	0.4				ug/L	27	Standard
	Y	89	687710.0	0.9				ug/L	383780	Standard
>	**Rh**	103	48.3	11.9				ug/L	18	Standard
	Mo	98	11816.3	1.1	**4.2765**	0.072	1.7	ug/L	31	Standard
	Ag	107	1411.4	4.5	**0.1899**	0.008	4.0	ug/L	84	Standard
	Cd	111	1971.0	1.8	**0.8883**	0.019	2.1	mg/L	4	Standard
	Cd	114	4664.8	1.8	**0.8659**	0.021	2.5	ug/L	15	Standard
>	**In**	115	534504.1	1.2				ug/L	515400	Standard
	Sn	118	9728.1	2.5	**1.4395**	0.057	3.9	ug/L	1043	Standard
	Sb	123	5858.1	3.7	**1.2180**	0.055	4.5	ug/L	227	Standard
	Ba	135	343595.4	0.9	**164.8188**	3.343	2.0	ug/L	31	Standard
	Ce	140	904246.5	1.5				ug/L	120	Standard
>	**Tb**	159	929077.6	1.4				ug/L	785817	Standard
	Ho	165	20382.0	1.9				ug/L	7	Standard
	Tl	203	8751.5	1.5	**0.9686**	0.028	2.9	ug/L	6	Standard
	Tl	205	7040.0	1.3	**0.8909**	0.012	1.3	ug/L	8	Standard
	Pb	206	269096.1	1.0	**48.0135**	0.893	1.9	ug/L	201	Standard
	Pb	207	224906.7	1.1	**44.3862**	1.110	2.5	ug/L	177	Standard
	Pb	208	846557.7	1.0	**41.5955**	1.382	3.3	ug/L	722	Standard
	U	238	16518.0	1.0	**2.1348**	0.065	3.0	ug/L	8	Standard
>	**Bi**	209	500556.2	2.5				ug/L	444563	Standard

Sample ID: L1606145105
Report Date/Time: Tuesday, July 05, 2016 11:02:30
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**13.2844**	23.001	173.1	mg/L	0	Standard
Mg	24	855.0	4.6	**0.6981**	0.024	3.5	mg/L	38	Standard
K	39	1165.0	4.5	**3.3481**	0.167	5.0	mg/L	33	Standard
Ca	43	118.3	2.4	**6.8217**	0.520	7.6	mg/L	48	Standard
Fe	54	77319.8	1.6	**42.3823**	1.246	2.9	mg/L	131	Standard
Fe	57	19831.3	1.4	**41.4206**	1.327	3.2	mg/L	153	Standard
Sc-1	45	39190.6	1.9				mg/L	26715	Standard
Cl	35	125192.6	2.4				ug/L	102686	Standard
Kr	83	2.3	89.2				ug/L	2	Standard
Br	81	3120.3	1.7				ug/L	1897	Standard
P	31	27561.7	1.5				ug/L	18491	Standard
S	34	6189.6	2.3				ug/L	3734	Standard
Sr	88	120.0	20.8				ug/L	47	Standard
C	12	446.7	27.6				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	183.3	22.0				mg/L	23	Standard
Dy	164	32276.2	0.2				mg/L	23	Standard
Ho-1	165	20382.0	1.9				mg/L	7	Standard
Er	166	19766.2	2.6				mg/L	13	Standard
I	127	39155.6	4.6				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		182.523	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		115.510	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145105
Report Date/Time: Tuesday, July 05, 2016 11:02:30
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.707
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		112.595
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Ti 47 Upper, S, EEE	Ti	47	
Cr 53 Upper, S, EEE	Cr	53	

Sample ID: L1606145105
Report Date/Time: Tuesday, July 05, 2016 11:02:30
Page 3

Approved: July 06, 2016

Mn 55 Upper, S, EEE	Mn	55
Zn 66 Upper, S, EEE	Zn	66
Se-1 77 Upper, S, EEE	Se-1	77
Ba 135 Upper, S, EEE	Ba	135

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145107
Sample Date/Time: Tuesday, July 05, 2016 11:03:24
Number of Replicates: 3
Autosampler Position: 218
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	109709.1	1.5				ug/L	59517	Standard
	Be	9	3635.4	4.4	**2.2384**	0.131	5.8	ug/L	10	Standard
	Al	27	7905627.2	2.0	**28.1821**	0.827	2.9	ug/L	1872	Standard
	Sc	45	37616.6	1.4				ug/L	26715	Standard
	Ti	47	69388.6	1.8	**240.1521**	3.763	1.6	ug/L	54	Standard
	V	51	342535.6	1.2	**38.9308**	0.728	1.9	ug/L	1419	Standard
	Cr	52	390616.2	0.7	**44.2829**	0.338	0.8	ug/L	7059	Standard
	Cr	53	194072.8	8.7	**179.4677**	16.874	9.4	ug/L	398	Standard
	Mn	55	1910553.9	1.2	**248.5998**	4.612	1.9	ug/L	824	Standard
	Co	59	51087.0	1.5	**6.1690**	0.115	1.9	ug/L	167	Standard
	Ni	60	52765.4	1.2	**23.0352**	0.256	1.1	ug/L	391	Standard
	Cu	65	52726.9	0.4	**23.0606**	0.120	0.5	ug/L	147	Standard
	Zn	66	190722.6	1.0	**142.1825**	0.962	0.7	ug/L	184	Standard
>	Ge	72	513489.2	0.8				ug/L	422998	Standard
	As	75	71231.2	0.5	**51.7747**	0.529	1.0	ug/L	-58	Standard
	Se	82	249.5	6.5	**1.5478**	0.127	8.2	ug/L	15	Standard
	Se-1	77	16454.6	7.9	**178.1501**	15.353	8.6	ug/L	58	Standard
>	Ga	71	24666.7	1.8				mg/L	113	Standard
	Rb	85	164243.8	1.0				ug/L	27	Standard
	Y	89	634729.6	2.2				ug/L	383780	Standard
>	Rh	103	63.3	18.2				ug/L	18	Standard
	Mo	98	18846.0	0.8	**6.4893**	0.189	2.9	ug/L	31	Standard
	Ag	107	937.0	2.7	**0.1155**	0.004	3.8	ug/L	84	Standard
	Cd	111	4443.5	1.2	**1.9084**	0.039	2.0	mg/L	4	Standard
	Cd	114	10909.6	4.9	**1.9304**	0.098	5.1	ug/L	15	Standard
>	In	115	562417.7	2.6				ug/L	515400	Standard
	Sn	118	8864.3	5.2	**1.2219**	0.059	4.9	ug/L	1043	Standard
	Sb	123	9482.0	2.9	**1.8820**	0.030	1.6	ug/L	227	Standard
	Ba	135	193348.9	0.6	**88.1570**	2.232	2.5	ug/L	31	Standard
	Ce	140	806506.2	0.9				ug/L	120	Standard
>	Tb	159	968120.2	0.9				ug/L	785817	Standard
	Ho	165	15471.2	1.3				ug/L	7	Standard
	Tl	203	20732.5	0.4	**2.2137**	0.018	0.8	ug/L	6	Standard
	Tl	205	17249.8	1.3	**2.1112**	0.041	1.9	ug/L	8	Standard
	Pb	206	299572.9	0.3	**51.4663**	0.481	0.9	ug/L	201	Standard
	Pb	207	250580.1	0.2	**47.6121**	0.366	0.8	ug/L	177	Standard
	Pb	208	967157.6	0.8	**45.7453**	0.329	0.7	ug/L	722	Standard
	U	238	14798.9	1.7	**1.8407**	0.030	1.6	ug/L	8	Standard
>	Bi	209	519770.6	0.7				ug/L	444563	Standard

Sample ID: L1606145107
Report Date/Time: Tuesday, July 05, 2016 11:05:41
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**13.9090**	24.082	173.1	mg/L	0	Standard
Mg	24	353.3	20.5	**0.2817**	0.067	23.7	mg/L	38	Standard
K	39	568.3	15.8	**1.6404**	0.302	18.4	mg/L	33	Standard
Ca	43	153.3	18.5	**11.2881**	3.405	30.2	mg/L	48	Standard
Fe	54	42628.9	0.4	**24.2898**	0.303	1.2	mg/L	131	Standard
Fe	57	10755.5	0.6	**23.1886**	0.357	1.5	mg/L	153	Standard
Sc-1	45	37616.6	1.4				mg/L	26715	Standard
Cl	35	132874.8	0.7				ug/L	102686	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	2566.9	8.8				ug/L	1897	Standard
P	31	28668.8	3.4				ug/L	18491	Standard
S	34	6544.7	2.9				ug/L	3734	Standard
Sr	88	140.0	9.4				ug/L	47	Standard
C	12	416.7	10.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	306.7	21.0				mg/L	23	Standard
Dy	164	23970.0	2.2				mg/L	23	Standard
Ho-1	165	15471.2	1.3				mg/L	7	Standard
Er	166	14964.1	1.6				mg/L	13	Standard
I	127	23219.5	5.4				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		184.333	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		121.393	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145107
Report Date/Time: Tuesday, July 05, 2016 11:05:41
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.123
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		116.917
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Ti 47 Upper, S, EEE	Ti	47	
Cr 53 Upper, S, EEE	Cr	53	

Sample ID: L1606145107
Report Date/Time: Tuesday, July 05, 2016 11:05:41
Page 3

Approved: July 06, 2016

Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ge 72 Int Std for sample	Ge	72	Rerun sample
Se-1 77 Upper, S, EEE	Se-1	77	

Sample ID: L1606145107
Report Date/Time: Tuesday, July 05, 2016 11:05:41
Page 4

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145108
Sample Date/Time: Tuesday, July 05, 2016 11:06:36
Number of Replicates: 3
Autosampler Position: 219
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	115300.4	0.2				ug/L	59517	Standard
	Be	9	6846.5	2.1	**4.0334**	0.075	1.9	ug/L	10	Standard
	Al	27	8044256.6	3.1	**27.2779**	0.787	2.9	ug/L	1872	Standard
	Sc	45	43786.5	0.7				ug/L	26715	Standard
	Ti	47	177713.0	1.7	**608.3997**	18.233	3.0	ug/L	54	Standard
	V	51	562128.8	0.9	**63.2529**	0.893	1.4	ug/L	1419	Standard
	Cr	52	668635.3	1.5	**75.5984**	1.722	2.3	ug/L	7059	Standard
	Cr	53	233924.5	2.8	**213.7912**	6.209	2.9	ug/L	398	Standard
	Mn	55	1784049.0	0.8	**229.4141**	3.151	1.4	ug/L	824	Standard
	Co	59	94382.8	1.3	**11.2859**	0.270	2.4	ug/L	167	Standard
	Ni	60	90035.4	0.6	**38.9880**	0.661	1.7	ug/L	391	Standard
	Cu	65	75578.4	1.1	**32.7069**	0.789	2.4	ug/L	147	Standard
	Zn	66	263196.7	0.4	**193.9978**	2.967	1.5	ug/L	184	Standard
>	Ge	72	519575.6	1.4				ug/L	422998	Standard
	As	75	71823.8	0.7	**51.6006**	0.994	1.9	ug/L	-58	Standard
	Se	82	246.5	4.2	**1.5052**	0.047	3.1	ug/L	15	Standard
	Se-1	77	19182.2	5.3	**205.2715**	10.844	5.3	ug/L	58	Standard
>	Ga	71	40878.4	0.4				mg/L	113	Standard
	Rb	85	376160.7	0.9				ug/L	27	Standard
	Y	89	769310.7	0.4				ug/L	383780	Standard
>	Rh	103	66.7	15.6				ug/L	18	Standard
	Mo	98	18968.5	1.1	**6.5743**	0.126	1.9	ug/L	31	Standard
	Ag	107	1476.7	1.7	**0.1902**	0.005	2.7	ug/L	84	Standard
	Cd	111	4714.0	0.8	**2.0384**	0.033	1.6	mg/L	4	Standard
	Cd	114	11218.9	3.2	**1.9988**	0.074	3.7	ug/L	15	Standard
>	In	115	558549.2	0.9				ug/L	515400	Standard
	Sn	118	11075.7	4.1	**1.5838**	0.058	3.7	ug/L	1043	Standard
	Sb	123	9782.9	0.8	**1.9557**	0.004	0.2	ug/L	227	Standard
	Ba	135	356713.7	1.4	**163.7164**	1.201	0.7	ug/L	31	Standard
	Ce	140	1269887.9	0.3				ug/L	120	Standard
>	Tb	159	974069.9	0.7				ug/L	785817	Standard
	Ho	165	26974.0	1.7				ug/L	7	Standard
	Tl	203	25905.1	1.0	**2.7627**	0.008	0.3	ug/L	6	Standard
	Tl	205	21091.3	2.3	**2.5788**	0.054	2.1	ug/L	8	Standard
	Pb	206	399454.5	0.5	**68.5395**	0.825	1.2	ug/L	201	Standard
	Pb	207	334755.6	0.1	**63.5277**	0.561	0.9	ug/L	177	Standard
	Pb	208	1275635.6	0.2	**60.2624**	0.374	0.6	ug/L	722	Standard
	U	238	26668.1	0.9	**3.3141**	0.030	0.9	ug/L	8	Standard
>	Bi	209	520549.4	0.8				ug/L	444563	Standard

Sample ID: L1606145108
Report Date/Time: Tuesday, July 05, 2016 11:08:53
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**11.9257**	20.647	173.1	mg/L	0	Standard
Mg	24	778.4	9.2	**0.5628**	0.056	10.0	mg/L	38	Standard
K	39	1511.7	6.0	**3.9097**	0.271	6.9	mg/L	33	Standard
Ca	43	190.0	22.9	**12.3692**	4.209	34.0	mg/L	48	Standard
Fe	54	91363.6	2.2	**44.8185**	1.085	2.4	mg/L	131	Standard
Fe	57	24494.8	2.4	**45.8301**	1.415	3.1	mg/L	153	Standard
Sc-1	45	43786.5	0.7				mg/L	26715	Standard
Cl	35	129997.3	0.9				ug/L	102686	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	2873.6	6.2				ug/L	1897	Standard
P	31	27894.0	1.7				ug/L	18491	Standard
S	34	6764.8	2.2				ug/L	3734	Standard
Sr	88	176.7	8.6				ug/L	47	Standard
C	12	480.0	3.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	253.3	11.4				mg/L	23	Standard
Dy	164	41874.4	2.0				mg/L	23	Standard
Ho-1	165	26974.0	1.7				mg/L	7	Standard
Er	166	26029.0	1.2				mg/L	13	Standard
I	127	24075.8	5.6				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		193.728	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		122.832	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145108
Report Date/Time: Tuesday, July 05, 2016 11:08:53
Page 2

Approved: July 06, 2016

	Rb	85	
	Y	89	
>	Rh	103	
	Mo	98	
	Ag	107	
	Cd	111	
	Cd	114	
>	In	115	108.372
	Sn	118	
	Sb	123	
	Ba	135	
	Ce	140	
>	Tb	159	
	Ho	165	
	Tl	203	
	Tl	205	
	Pb	206	
	Pb	207	
	Pb	208	
	U	238	
>	Bi	209	117.092
	Na	23	
	Mg	24	
	K	39	
	Ca	43	
	Fe	54	
	Fe	57	
>	Sc-1	45	
	Cl	35	
	Kr	83	
	Br	81	
	P	31	
	S	34	
	Sr	88	
	C	12	
	N	14	
	Hg	202	
	Dy	164	
	Ho-1	165	
	Er	166	
	I	127	

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Ti 47 Upper, S, EEE	Ti	47	
Cr 53 Upper, S, EEE	Cr	53	

Sample ID: L1606145108
Report Date/Time: Tuesday, July 05, 2016 11:08:53
Page 3

Approved: July 06, 2016

Mn 55 Upper, S, EEE	Mn	55	
Zn 66 Upper, S, EEE	Zn	66	
Ge 72 Int Std for sample	Ge	72	Rerun sample
Se-1 77 Upper, S, EEE	Se-1	77	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606145108
Report Date/Time: Tuesday, July 05, 2016 11:08:53
Page 4

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606144503
Sample Date/Time: Tuesday, July 05, 2016 11:09:48
Number of Replicates: 3
Autosampler Position: 220
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	58967.9	1.9				ug/L	59517	Standard
	Be	9	735.0	5.3	**0.8229**	0.062	7.6	ug/L	10	Standard
	Al	27	8232717.6	1.5	**54.6151**	1.353	2.5	ug/L	1872	Standard
	Sc	45	35428.0	1.7				ug/L	26715	Standard
	Ti	47	5309.9	4.9	**19.4712**	0.822	4.2	ug/L	54	Standard
	V	51	177759.2	1.9	**21.5762**	0.278	1.3	ug/L	1419	Standard
	Cr	52	136308.2	1.7	**15.9537**	0.088	0.6	ug/L	7059	Standard
	Cr	53	31922.0	7.2	**31.2822**	2.011	6.4	ug/L	398	Standard
	Mn	55	3732711.4	1.4	**520.8023**	7.726	1.5	ug/L	824	Standard
	Co	59	63826.5	0.6	**8.2713**	0.128	1.5	ug/L	167	Standard
	Ni	60	35407.0	2.7	**16.5140**	0.389	2.4	ug/L	391	Standard
	Cu	65	27552.0	1.2	**12.8834**	0.152	1.2	ug/L	147	Standard
	Zn	66	68317.0	1.3	**54.5023**	0.739	1.4	ug/L	184	Standard
>	Ge	72	478996.0	1.2				ug/L	422998	Standard
	As	75	5507.3	2.6	**4.3379**	0.061	1.4	ug/L	-58	Standard
	Se	82	121.2	14.5	**0.7065**	0.126	17.8	ug/L	15	Standard
	Se-1	77	1211.4	12.1	**13.4294**	1.589	11.8	ug/L	58	Standard
>	Ga	71	17138.0	3.8				mg/L	113	Standard
	Rb	85	215157.6	2.6				ug/L	27	Standard
	Y	89	624025.0	0.8				ug/L	383780	Standard
>	Rh	103	36.7	7.9				ug/L	18	Standard
	Mo	98	1681.9	2.9	**0.5857**	0.019	3.2	ug/L	31	Standard
	Ag	107	636.7	2.8	**0.0768**	0.003	3.9	ug/L	84	Standard
	Cd	111	2696.7	0.8	**1.1861**	0.014	1.2	mg/L	4	Standard
	Cd	114	6661.2	1.4	**1.2069**	0.011	0.9	ug/L	15	Standard
>	In	115	548315.1	0.7				ug/L	515400	Standard
	Sn	118	1376.7	21.9	**0.0427**	0.050	117.8	ug/L	1043	Standard
	Sb	123	763.4	20.5	**0.1405**	0.033	23.4	ug/L	227	Standard
	Ba	135	486005.6	0.2	**227.2373**	1.036	0.5	ug/L	31	Standard
	Ce	140	1188503.1	0.8				ug/L	120	Standard
>	Tb	159	857273.9	0.6				ug/L	785817	Standard
	Ho	165	16704.2	0.6				ug/L	7	Standard
	Tl	203	934.4	4.2	**0.1071**	0.005	4.5	ug/L	6	Standard
	Tl	205	773.4	8.3	**0.0984**	0.009	8.7	ug/L	8	Standard
	Pb	206	181321.2	0.4	**34.4939**	0.134	0.4	ug/L	201	Standard
	Pb	207	149581.4	0.5	**31.4697**	0.185	0.6	ug/L	177	Standard
	Pb	208	642010.8	0.6	**33.6279**	0.141	0.4	ug/L	722	Standard
	U	238	4887.1	1.4	**0.6719**	0.010	1.4	ug/L	8	Standard
>	Bi	209	469171.3	0.1				ug/L	444563	Standard

Sample ID: L1606144503
Report Date/Time: Tuesday, July 05, 2016 11:12:05
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	676.7	5.7	**0.6075**	0.045	7.3	mg/L	38	Standard
K	39	433.3	12.0	**1.3035**	0.195	15.0	mg/L	33	Standard
Ca	43	71.7	28.2	**2.6014**	2.240	86.1	mg/L	48	Standard
Fe	54	18412.4	3.1	**11.0733**	0.150	1.4	mg/L	131	Standard
Fe	57	4905.8	1.8	**10.9803**	0.097	0.9	mg/L	153	Standard
> **Sc-1**	45	35428.0	1.7				mg/L	26715	Standard
Cl	35	115274.9	1.1				ug/L	102686	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	11237.5	4.2				ug/L	1897	Standard
P	31	20882.7	1.3				ug/L	18491	Standard
S	34	4217.3	1.1				ug/L	3734	Standard
Sr	88	130.0	43.3				ug/L	47	Standard
C	12	230.0	19.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	103.3	20.1				mg/L	23	Standard
Dy	164	28748.7	1.3				mg/L	23	Standard
Ho-1	165	16704.2	0.6				mg/L	7	Standard
Er	166	15588.0	1.5				mg/L	13	Standard
I	127	43800.0	5.4				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		99.078	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.238	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144503
Report Date/Time: Tuesday, July 05, 2016 11:12:05
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.386
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.535
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606144503
Report Date/Time: Tuesday, July 05, 2016 11:12:05
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606144507
Sample Date/Time: Tuesday, July 05, 2016 11:12:59
Number of Replicates: 3
Autosampler Position: 221
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	61389.1	1.3				ug/L	59517	Standard
	Be	9	801.7	3.5	**0.8629**	0.035	4.1	ug/L	10	Standard
	Al	27	16239193.8	1.3	**103.4670**	0.803	0.8	ug/L	1872	Standard
	Sc	45	33152.9	3.9				ug/L	26715	Standard
	Ti	47	3309.7	4.9	**12.7358**	0.436	3.4	ug/L	54	Standard
	V	51	154376.9	1.5	**19.7873**	0.570	2.9	ug/L	1419	Standard
	Cr	52	182451.6	0.6	**22.9734**	0.662	2.9	ug/L	7059	Standard
	Cr	53	30803.0	5.1	**31.9028**	1.278	4.0	ug/L	398	Standard
	Mn	55	3250036.0	0.9	**479.1339**	10.524	2.2	ug/L	824	Standard
	Co	59	64709.9	0.9	**8.8615**	0.150	1.7	ug/L	167	Standard
	Ni	60	45396.2	0.8	**22.4433**	0.365	1.6	ug/L	391	Standard
	Cu	65	45400.6	1.0	**22.4890**	0.351	1.6	ug/L	147	Standard
	Zn	66	103172.4	0.9	**87.0675**	1.942	2.2	ug/L	184	Standard
>	Ge	72	453421.2	2.2				ug/L	422998	Standard
	As	75	13039.3	1.1	**10.7771**	0.263	2.4	ug/L	-58	Standard
	Se	82	76.6	11.2	**0.4031**	0.068	16.8	ug/L	15	Standard
	Se-1	77	615.7	11.8	**6.9006**	0.798	11.6	ug/L	58	Standard
>	Ga	71	18160.9	1.8				mg/L	113	Standard
	Rb	85	189390.0	2.2				ug/L	27	Standard
	Y	89	636710.9	1.2				ug/L	383780	Standard
>	Rh	103	33.3	37.7				ug/L	18	Standard
	Mo	98	518.6	4.6	**0.1814**	0.008	4.7	ug/L	31	Standard
	Ag	107	598.0	13.5	**0.0746**	0.012	15.4	ug/L	84	Standard
	Cd	111	1869.2	2.7	**0.8523**	0.022	2.6	mg/L	4	Standard
	Cd	114	4708.0	1.7	**0.8844**	0.008	0.9	ug/L	15	Standard
>	In	115	528155.1	0.8				ug/L	515400	Standard
	Sn	118	1220.1	19.5	**0.0245**	0.039	159.2	ug/L	1043	Standard
	Sb	123	1538.9	3.7	**0.3117**	0.012	3.7	ug/L	227	Standard
	Ba	135	375687.6	0.3	**182.3635**	1.875	1.0	ug/L	31	Standard
	Ce	140	1225726.9	0.7				ug/L	120	Standard
>	Tb	159	833959.3	0.9				ug/L	785817	Standard
	Ho	165	19961.5	2.9				ug/L	7	Standard
	Tl	203	1398.7	7.3	**0.1660**	0.011	6.5	ug/L	6	Standard
	Tl	205	1258.4	2.0	**0.1685**	0.005	3.1	ug/L	8	Standard
	Pb	206	525459.7	0.9	**102.4626**	1.464	1.4	ug/L	201	Standard
	Pb	207	445735.9	0.6	**96.1318**	0.540	0.6	ug/L	177	Standard
	Pb	208	1863328.7	0.5	**100.0487**	1.095	1.1	ug/L	722	Standard
	U	238	3947.2	5.3	**0.5551**	0.024	4.4	ug/L	8	Standard
>	Bi	209	458154.8	1.0				ug/L	444563	Standard

Sample ID: L1606144507
Report Date/Time: Tuesday, July 05, 2016 11:15:16
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	113.3	12.7	**0.0814**	0.018	22.3	mg/L	38	Standard
K	39	460.0	10.0	**1.4976**	0.203	13.6	mg/L	33	Standard
Ca	43	71.7	29.0	**3.1501**	2.246	71.3	mg/L	48	Standard
Fe	54	16348.2	2.7	**10.5068**	0.225	2.1	mg/L	131	Standard
Fe	57	4519.0	5.9	**10.7959**	0.248	2.3	mg/L	153	Standard
Sc-1	45	33152.9	3.9				mg/L	26715	Standard
Cl	35	115120.4	1.5				ug/L	102686	Standard
Kr	83	3.3	45.8				ug/L	2	Standard
Br	81	3303.7	5.2				ug/L	1897	Standard
P	31	19818.0	2.4				ug/L	18491	Standard
S	34	4165.6	2.5				ug/L	3734	Standard
Sr	88	96.7	35.2				ug/L	47	Standard
C	12	176.7	16.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	163.3	19.7				mg/L	23	Standard
Dy	164	34130.8	0.6				mg/L	23	Standard
Ho-1	165	19961.5	2.9				mg/L	7	Standard
Er	166	17316.6	3.2				mg/L	13	Standard
I	127	24820.3	3.3				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		103.146	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		107.192	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144507
Report Date/Time: Tuesday, July 05, 2016 11:15:16
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.475
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.057
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606144507
Report Date/Time: Tuesday, July 05, 2016 11:15:16
Page 3

Approved: July 06, 2016

Pb 206 Upper, S, EEE	Pb	206
Pb 208 Upper, S, EEE	Pb	208

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606144508
Sample Date/Time: Tuesday, July 05, 2016 11:16:11
Number of Replicates: 3
Autosampler Position: 222
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	64903.7	0.9				ug/L	59517	Standard
	Be	9	825.0	11.8	**0.8387**	0.099	11.8	ug/L	10	Standard
	Al	27	23443690.4	1.1	**141.2975**	2.619	1.9	ug/L	1872	Standard
	Sc	45	34385.7	2.0				ug/L	26715	Standard
	Ti	47	4006.5	1.7	**14.9742**	0.331	2.2	ug/L	54	Standard
	V	51	161428.6	1.6	**20.0296**	0.394	2.0	ug/L	1419	Standard
	Cr	52	127935.9	0.6	**15.2771**	0.113	0.7	ug/L	7059	Standard
	Cr	53	19500.9	1.4	**19.4010**	0.396	2.0	ug/L	398	Standard
	Mn	55	3729725.9	0.2	**532.2377**	3.503	0.7	ug/L	824	Standard
	Co	59	67689.2	0.4	**8.9734**	0.082	0.9	ug/L	167	Standard
	Ni	60	45008.4	0.7	**21.5312**	0.145	0.7	ug/L	391	Standard
	Cu	65	24847.7	1.2	**11.8778**	0.197	1.7	ug/L	147	Standard
	Zn	66	45863.0	0.8	**37.3734**	0.528	1.4	ug/L	184	Standard
>	**Ge**	72	468315.2	0.6				ug/L	422998	Standard
	As	75	4552.8	1.0	**3.6764**	0.042	1.1	ug/L	-58	Standard
	Se	82	52.7	20.8	**0.1995**	0.082	41.3	ug/L	15	Standard
	Se-1	77	367.7	2.8	**3.7114**	0.142	3.8	ug/L	58	Standard
>	**Ga**	71	18396.2	3.7				mg/L	113	Standard
	Rb	85	184872.8	1.2				ug/L	27	Standard
	Y	89	657495.0	1.5				ug/L	383780	Standard
>	**Rh**	103	23.3	107.9				ug/L	18	Standard
	Mo	98	421.2	2.4	**0.1440**	0.004	2.8	ug/L	31	Standard
	Ag	107	473.7	4.3	**0.0559**	0.004	7.1	ug/L	84	Standard
	Cd	111	178.9	4.3	**0.0769**	0.004	5.1	mg/L	4	Standard
	Cd	114	422.3	1.9	**0.0743**	0.003	3.6	ug/L	15	Standard
>	**In**	115	534021.2	1.6				ug/L	515400	Standard
	Sn	118	1095.0	20.4	**0.0013**	0.035	2649.2	ug/L	1043	Standard
	Sb	123	466.4	31.4	**0.0817**	0.030	36.1	ug/L	227	Standard
	Ba	135	398222.1	0.7	**191.1971**	2.894	1.5	ug/L	31	Standard
	Ce	140	1354005.7	1.4				ug/L	120	Standard
>	**Tb**	159	854017.5	2.4				ug/L	785817	Standard
	Ho	165	21572.0	2.2				ug/L	7	Standard
	Tl	203	941.4	4.7	**0.1100**	0.008	7.0	ug/L	6	Standard
	Tl	205	776.7	3.5	**0.1008**	0.006	5.9	ug/L	8	Standard
	Pb	206	58153.4	1.0	**11.2275**	0.288	2.6	ug/L	201	Standard
	Pb	207	47609.7	0.2	**10.1602**	0.203	2.0	ug/L	177	Standard
	Pb	208	201890.5	1.1	**10.7270**	0.171	1.6	ug/L	722	Standard
	U	238	3940.5	1.1	**0.5508**	0.015	2.6	ug/L	8	Standard
>	**Bi**	209	461184.9	2.1				ug/L	444563	Standard

Sample ID: L1606144508
Report Date/Time: Tuesday, July 05, 2016 11:18:27
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**15.5379**	26.904	173.1	mg/L	0	Standard
Mg	24	121.7	4.7	**0.0850**	0.003	3.8	mg/L	38	Standard
K	39	315.0	16.5	**0.9413**	0.167	17.7	mg/L	33	Standard
Ca	43	193.3	4.0	**17.7553**	0.524	3.0	mg/L	48	Standard
Fe	54	18303.5	2.1	**11.3469**	0.178	1.6	mg/L	131	Standard
Fe	57	4834.1	0.4	**11.1587**	0.275	2.5	mg/L	153	Standard
Sc-1	45	34385.7	2.0				mg/L	26715	Standard
Cl	35	119520.9	0.7				ug/L	102686	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	2616.9	5.8				ug/L	1897	Standard
P	31	20876.0	2.0				ug/L	18491	Standard
S	34	4800.8	3.6				ug/L	3734	Standard
Sr	88	108.3	17.5				ug/L	47	Standard
C	12	63.3	36.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	30.0	57.7				mg/L	23	Standard
Dy	164	36615.9	1.3				mg/L	23	Standard
Ho-1	165	21572.0	2.2				mg/L	7	Standard
Er	166	18658.1	0.8				mg/L	13	Standard
I	127	15055.8	3.2				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		109.051	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.713	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144508
Report Date/Time: Tuesday, July 05, 2016 11:18:27
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.613
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		103.739
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606144508
Report Date/Time: Tuesday, July 05, 2016 11:18:27
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606144509
Sample Date/Time: Tuesday, July 05, 2016 11:19:22
Number of Replicates: 3
Autosampler Position: 223
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	65787.5	0.7				ug/L	59517	Standard
	Be	9	766.7	15.6	**0.7672**	0.129	16.8	ug/L	10	Standard
	Al	27	21794353.3	0.8	**129.5855**	1.775	1.4	ug/L	1872	Standard
	Sc	45	34634.5	1.6				ug/L	26715	Standard
	Ti	47	3670.1	1.4	**13.7771**	0.357	2.6	ug/L	54	Standard
	V	51	118166.5	2.8	**14.6961**	0.323	2.2	ug/L	1419	Standard
	Cr	52	108824.5	1.7	**12.9319**	0.457	3.5	ug/L	7059	Standard
	Cr	53	15558.0	2.2	**15.4900**	0.610	3.9	ug/L	398	Standard
	Mn	55	2951295.7	0.2	**423.5613**	6.271	1.5	ug/L	824	Standard
	Co	59	53986.3	0.5	**7.1934**	0.139	1.9	ug/L	167	Standard
	Ni	60	38514.6	0.8	**18.5039**	0.332	1.8	ug/L	391	Standard
	Cu	65	19658.7	0.7	**9.4358**	0.198	2.1	ug/L	147	Standard
	Zn	66	37941.1	1.0	**31.0660**	0.222	0.7	ug/L	184	Standard
>	Ge	72	465689.2	1.7				ug/L	422998	Standard
	As	75	2790.3	2.8	**2.2853**	0.026	1.1	ug/L	-58	Standard
	Se	82	47.9	14.3	**0.1646**	0.047	28.5	ug/L	15	Standard
	Se-1	77	281.7	4.7	**2.7044**	0.105	3.9	ug/L	58	Standard
>	Ga	71	15155.9	4.1				mg/L	113	Standard
	Rb	85	146035.6	2.2				ug/L	27	Standard
	Y	89	640430.1	0.5				ug/L	383780	Standard
>	Rh	103	13.3	114.6				ug/L	18	Standard
	Mo	98	435.9	5.3	**0.1454**	0.008	5.4	ug/L	31	Standard
	Ag	107	423.3	4.0	**0.0472**	0.003	6.1	ug/L	84	Standard
	Cd	111	141.5	5.8	**0.0584**	0.003	5.5	mg/L	4	Standard
	Cd	114	357.3	12.2	**0.0605**	0.008	12.7	ug/L	15	Standard
>	In	115	547325.2	0.9				ug/L	515400	Standard
	Sn	118	1031.7	23.6	**-0.0131**	0.040	306.0	ug/L	1043	Standard
	Sb	123	468.3	21.2	**0.0800**	0.020	25.5	ug/L	227	Standard
	Ba	135	313028.8	1.1	**146.6280**	2.686	1.8	ug/L	31	Standard
	Ce	140	1077990.0	0.6				ug/L	120	Standard
>	Tb	159	866335.0	0.7				ug/L	785817	Standard
	Ho	165	18498.0	3.4				ug/L	7	Standard
	Tl	203	739.0	2.9	**0.0840**	0.002	2.3	ug/L	6	Standard
	Tl	205	630.0	4.4	**0.0789**	0.003	4.2	ug/L	8	Standard
	Pb	206	48931.2	1.6	**9.2769**	0.123	1.3	ug/L	201	Standard
	Pb	207	40058.9	0.6	**8.3945**	0.072	0.9	ug/L	177	Standard
	Pb	208	170404.7	0.4	**8.8920**	0.075	0.8	ug/L	722	Standard
	U	238	3169.3	0.8	**0.4349**	0.007	1.5	ug/L	8	Standard
>	Bi	209	469087.4	0.9				ug/L	444563	Standard

Sample ID: L1606144509
Report Date/Time: Tuesday, July 05, 2016 11:21:39
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**14.8872**	25.777	173.1	mg/L	0	Standard
Mg	24	100.0	8.7	**0.0633**	0.010	15.8	mg/L	38	Standard
K	39	288.3	8.9	**0.8451**	0.097	11.5	mg/L	33	Standard
Ca	43	228.3	7.0	**21.8553**	2.204	10.1	mg/L	48	Standard
Fe	54	12940.1	2.9	**7.9291**	0.195	2.5	mg/L	131	Standard
Fe	57	3518.7	6.8	**7.9289**	0.582	7.3	mg/L	153	Standard
> **Sc-1**	45	34634.5	1.6				mg/L	26715	Standard
Cl	35	121307.7	0.9				ug/L	102686	Standard
Kr	83	2.3	107.9				ug/L	2	Standard
Br	81	2903.6	6.8				ug/L	1897	Standard
P	31	22823.9	6.4				ug/L	18491	Standard
S	34	4914.1	1.7				ug/L	3734	Standard
Sr	88	120.0	4.2				ug/L	47	Standard
C	12	103.3	20.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	53.3	10.8				mg/L	23	Standard
Dy	164	30605.6	2.0				mg/L	23	Standard
Ho-1	165	18498.0	3.4				mg/L	7	Standard
Er	166	16288.8	5.4				mg/L	13	Standard
I	127	15886.7	2.1				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		110.536	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		110.093	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144509
Report Date/Time: Tuesday, July 05, 2016 11:21:39
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.194
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.517
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Al 27 Upper, S, EEE	Al	27	
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606144509
Report Date/Time: Tuesday, July 05, 2016 11:21:39
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606144510
Sample Date/Time: Tuesday, July 05, 2016 11:22:34
Number of Replicates: 3
Autosampler Position: 224
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	67285.8	1.4				ug/L	59517	Standard
	Be	9	843.4	7.9	**0.8264**	0.056	6.8	ug/L	10	Standard
	Al	27	10707433.3	0.6	**62.2417**	0.536	0.9	ug/L	1872	Standard
	Sc	45	34009.8	1.5				ug/L	26715	Standard
	Ti	47	1897.1	4.0	**6.8476**	0.267	3.9	ug/L	54	Standard
	V	51	117452.5	1.9	**14.2941**	0.209	1.5	ug/L	1419	Standard
	Cr	52	118454.6	1.4	**13.8375**	0.081	0.6	ug/L	7059	Standard
	Cr	53	17269.8	2.1	**16.8637**	0.557	3.3	ug/L	398	Standard
	Mn	55	2195267.9	0.8	**308.3614**	5.595	1.8	ug/L	824	Standard
	Co	59	52932.6	0.2	**6.9030**	0.145	2.1	ug/L	167	Standard
	Ni	60	34129.4	1.7	**16.0212**	0.170	1.1	ug/L	391	Standard
	Cu	65	23297.6	1.1	**10.9570**	0.125	1.1	ug/L	147	Standard
	Zn	66	52962.1	1.1	**42.5081**	0.637	1.5	ug/L	184	Standard
>	Ge	72	475774.6	2.0				ug/L	422998	Standard
	As	75	5685.9	1.4	**4.5079**	0.046	1.0	ug/L	-58	Standard
	Se	82	64.1	10.5	**0.2807**	0.056	19.9	ug/L	15	Standard
	Se-1	77	294.7	9.5	**2.7856**	0.306	11.0	ug/L	58	Standard
>	Ga	71	14221.7	0.8				mg/L	113	Standard
	Rb	85	144167.6	0.3				ug/L	27	Standard
	Y	89	628491.7	3.5				ug/L	383780	Standard
>	Rh	103	33.3	17.3				ug/L	18	Standard
	Mo	98	776.2	4.0	**0.2609**	0.006	2.5	ug/L	31	Standard
	Ag	107	453.7	0.9	**0.0503**	0.001	1.9	ug/L	84	Standard
	Cd	111	445.2	5.4	**0.1890**	0.006	3.4	mg/L	4	Standard
	Cd	114	1054.3	3.5	**0.1840**	0.003	1.4	ug/L	15	Standard
>	In	115	557481.8	2.3				ug/L	515400	Standard
	Sn	118	1206.7	15.4	**0.0122**	0.034	277.1	ug/L	1043	Standard
	Sb	123	933.4	7.3	**0.1724**	0.018	10.2	ug/L	227	Standard
	Ba	135	291889.8	0.6	**134.2623**	2.822	2.1	ug/L	31	Standard
	Ce	140	1070488.0	1.6				ug/L	120	Standard
>	Tb	159	874195.5	1.1				ug/L	785817	Standard
	Ho	165	17550.2	0.3				ug/L	7	Standard
	Tl	203	711.0	2.4	**0.0784**	0.001	1.7	ug/L	6	Standard
	Tl	205	633.3	4.3	**0.0771**	0.003	4.1	ug/L	8	Standard
	Pb	206	222719.4	0.9	**41.2710**	0.449	1.1	ug/L	201	Standard
	Pb	207	185657.2	0.3	**38.0499**	0.491	1.3	ug/L	177	Standard
	Pb	208	785337.8	0.7	**40.0697**	0.390	1.0	ug/L	722	Standard
	U	238	3434.4	1.7	**0.4590**	0.005	1.1	ug/L	8	Standard
>	Bi	209	481789.4	1.1				ug/L	444563	Standard

Sample ID: L1606144510
Report Date/Time: Tuesday, July 05, 2016 11:24:50
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	61.7	24.8	**0.0271**	0.015	54.1	mg/L	38	Standard
K	39	261.7	14.3	**0.7713**	0.137	17.8	mg/L	33	Standard
Ca	43	70.0	37.8	**2.8074**	3.379	120.4	mg/L	48	Standard
Fe	54	12695.2	1.8	**7.9239**	0.243	3.1	mg/L	131	Standard
Fe	57	3348.7	2.0	**7.6685**	0.126	1.6	mg/L	153	Standard
> **Sc-1**	45	34009.8	1.5				mg/L	26715	Standard
Cl	35	112652.2	3.4				ug/L	102686	Standard
Kr	83	2.7	78.1				ug/L	2	Standard
Br	81	3233.7	0.9				ug/L	1897	Standard
P	31	19951.5	2.1				ug/L	18491	Standard
S	34	4202.3	4.9				ug/L	3734	Standard
Sr	88	98.3	19.3				ug/L	47	Standard
C	12	130.0	38.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	106.7	23.6				mg/L	23	Standard
Dy	164	30540.8	1.1				mg/L	23	Standard
Ho-1	165	17550.2	0.3				mg/L	7	Standard
Er	166	16245.4	1.0				mg/L	13	Standard
I	127	22099.4	2.8				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		113.054	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		112.477	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144510
Report Date/Time: Tuesday, July 05, 2016 11:24:50
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.165
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.374
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, July 05, 2016 11:25:46
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	62194.1	1.8				ug/L	59517	Standard
	Be	9	43385.4	0.9	**47.7293**	0.929	1.9	ug/L	10	Standard
	Al	27	7477094.9	1.1	**47.0246**	0.891	1.9	ug/L	1872	Standard
	Sc	45	29181.4	0.5				ug/L	26715	Standard
	Ti	47	25259.4	0.6	**96.0196**	0.868	0.9	ug/L	54	Standard
	V	51	376869.1	1.4	**47.1482**	0.392	0.8	ug/L	1419	Standard
	Cr	52	384616.4	0.9	**48.0425**	0.261	0.5	ug/L	7059	Standard
	Cr	53	49186.4	1.2	**49.7150**	0.294	0.6	ug/L	398	Standard
	Mn	55	329550.8	1.8	**47.0561**	0.547	1.2	ug/L	824	Standard
	Co	59	355041.7	2.0	**47.3232**	0.832	1.8	ug/L	167	Standard
	Ni	60	99023.6	1.6	**47.7629**	0.456	1.0	ug/L	391	Standard
	Cu	65	98958.8	1.7	**47.6895**	0.531	1.1	ug/L	147	Standard
	Zn	66	57970.7	1.3	**47.4325**	0.569	1.2	ug/L	184	Standard
>	Ge	72	466810.9	0.7				ug/L	422998	Standard
	As	75	58770.3	0.9	**46.9912**	0.219	0.5	ug/L	-58	Standard
	Se	82	6074.8	1.7	**46.7458**	0.794	1.7	ug/L	15	Standard
	Se-1	77	4177.6	1.2	**49.2498**	0.363	0.7	ug/L	58	Standard
>	Ga	71	68.3	22.4				mg/L	113	Standard
	Rb	85	1025.0	3.4				ug/L	27	Standard
	Y	89	407133.1	2.5				ug/L	383780	Standard
>	Rh	103	30.0	50.0				ug/L	18	Standard
	Mo	98	261587.0	1.0	**93.3540**	0.443	0.5	ug/L	31	Standard
	Ag	107	336725.6	1.2	**47.4083**	0.372	0.8	ug/L	84	Standard
	Cd	111	107984.0	1.0	**48.1146**	0.618	1.3	mg/L	4	Standard
	Cd	114	261565.4	1.4	**48.0258**	0.382	0.8	ug/L	15	Standard
>	In	115	543076.2	0.7				ug/L	515400	Standard
	Sn	118	288798.7	1.9	**47.1544**	0.692	1.5	ug/L	1043	Standard
	Sb	123	237278.6	1.2	**49.1762**	0.361	0.7	ug/L	227	Standard
	Ba	135	100307.0	1.3	**47.3315**	0.378	0.8	ug/L	31	Standard
	Ce	140	1143.4	23.0				ug/L	120	Standard
>	Tb	159	834060.6	0.6				ug/L	785817	Standard
	Ho	165	38.3	66.9				ug/L	7	Standard
	Tl	203	401949.2	0.6	**47.6074**	0.186	0.4	ug/L	6	Standard
	Tl	205	352686.4	1.3	**47.9509**	0.567	1.2	ug/L	8	Standard
	Pb	206	252819.9	0.4	**48.1023**	0.359	0.7	ug/L	201	Standard
	Pb	207	231007.5	1.0	**48.6132**	0.251	0.5	ug/L	177	Standard
	Pb	208	933186.5	0.6	**48.8922**	0.682	1.4	ug/L	722	Standard
	U	238	345826.5	1.6	**47.7023**	0.615	1.3	ug/L	8	Standard
>	Bi	209	469300.9	0.9				ug/L	444563	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 11:28:03
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**17.8640**	30.933	173.2	mg/L	0	Standard
Mg	24	4187.2	5.4	**4.7809**	0.274	5.7	mg/L	38	Standard
K	39	1096.7	4.6	**4.2663**	0.210	4.9	mg/L	33	Standard
Ca	43	81.7	15.4	**5.8840**	1.786	30.3	mg/L	48	Standard
Fe	54	6549.3	3.2	**4.7166**	0.160	3.4	mg/L	131	Standard
Fe	57	1873.4	8.2	**4.8332**	0.452	9.4	mg/L	153	Standard
> **Sc-1**	45	29181.4	0.5				mg/L	26715	Standard
Cl	35	119636.5	1.3				ug/L	102686	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1663.4	6.0				ug/L	1897	Standard
P	31	21340.0	3.4				ug/L	18491	Standard
S	34	4732.4	6.1				ug/L	3734	Standard
Sr	88	73.3	15.7				ug/L	47	Standard
C	12	30.0	88.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	0.0				mg/L	23	Standard
Dy	164	62.1	82.6				mg/L	23	Standard
Ho-1	165	38.3	66.9				mg/L	7	Standard
Er	166	26.7	21.7				mg/L	13	Standard
I	127	16135.3	1.1				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6			
Be	9	95.459		
Al	27	94.049		
Sc	45			
Ti	47	96.020		
V	51	94.296		
Cr	52	96.085		
Cr	53			
Mn	55	94.112		
Co	59	94.646		
Ni	60	95.526		
Cu	65	95.379		
Zn	66	94.865		
> Ge	72		110.358	
As	75	93.982		
Se	82	93.492		
Se-1	77			
> Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 11:28:03
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	93.354		
Ag	107	94.817		
Cd	111	96.229		
Cd	114			
> In	115			105.370
Sn	118	94.309		
Sb	123	98.352		
Ba	135	94.663		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	95.215		
Tl	205			
Pb	206			
Pb	207			
Pb	208	97.784		
U	238	95.405		
> Bi	209			105.565
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 11:28:03
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, July 05, 2016 11:28:57
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	63732.2	2.7				ug/L	59517	Standard
	Be	9	26.7	88.6	**-0.0024**	0.025	1040.4	ug/L	10	Standard
	Al	27	2688.6	20.4	**0.0040**	0.003	74.8	ug/L	1872	Standard
	Sc	45	29386.8	0.7				ug/L	26715	Standard
	Ti	47	34.3	8.9	**-0.1156**	0.010	8.7	ug/L	54	Standard
	V	51	1576.2	14.1	**0.0190**	0.026	137.2	ug/L	1419	Standard
	Cr	52	6976.9	2.0	**-0.0841**	0.007	8.6	ug/L	7059	Standard
	Cr	53	2505.2	1.2	**2.1511**	0.054	2.5	ug/L	398	Standard
	Mn	55	857.4	8.3	**-0.0078**	0.010	125.2	ug/L	824	Standard
	Co	59	217.3	4.9	**0.0040**	0.002	39.4	ug/L	167	Standard
	Ni	60	264.3	5.7	**-0.0714**	0.008	11.4	ug/L	391	Standard
	Cu	65	128.7	12.6	**-0.0167**	0.008	46.4	ug/L	147	Standard
	Zn	66	266.3	2.6	**0.0601**	0.007	11.2	ug/L	184	Standard
>	**Ge**	72	465388.3	1.2				ug/L	422998	Standard
	As	75	-51.1	37.1	**0.0106**	0.015	145.5	ug/L	-58	Standard
	Se	82	22.1	29.5	**-0.0350**	0.049	139.9	ug/L	15	Standard
	Se-1	77	178.7	5.4	**1.4737**	0.126	8.6	ug/L	58	Standard
>	**Ga**	71	40.0	50.0				mg/L	113	Standard
	Rb	85	43.3	52.0				ug/L	27	Standard
	Y	89	408149.9	1.3				ug/L	383780	Standard
>	**Rh**	103	11.7	24.7				ug/L	18	Standard
	Mo	98	203.4	7.9	**0.0631**	0.006	9.3	ug/L	31	Standard
	Ag	107	114.7	16.5	**0.0040**	0.003	64.7	ug/L	84	Standard
	Cd	111	14.6	11.7	**0.0023**	0.001	32.8	mg/L	4	Standard
	Cd	114	40.7	36.1	**0.0028**	0.003	94.1	ug/L	15	Standard
>	**In**	115	547595.3	0.4				ug/L	515400	Standard
	Sn	118	1465.1	25.8	**0.0571**	0.061	106.2	ug/L	1043	Standard
	Sb	123	769.9	30.2	**0.1419**	0.048	33.5	ug/L	227	Standard
	Ba	135	72.0	30.3	**0.0100**	0.010	102.5	ug/L	31	Standard
	Ce	140	465.0	17.8				ug/L	120	Standard
>	**Tb**	159	844940.1	1.4				ug/L	785817	Standard
	Ho	165	11.7	24.7				ug/L	7	Standard
	Tl	203	76.7	27.6	**0.0053**	0.002	47.5	ug/L	6	Standard
	Tl	205	68.3	48.7	**0.0023**	0.004	192.3	ug/L	8	Standard
	Pb	206	292.0	9.4	**0.0086**	0.005	60.7	ug/L	201	Standard
	Pb	207	248.7	4.3	**0.0032**	0.003	80.4	ug/L	177	Standard
	Pb	208	1010.7	9.0	**0.0031**	0.005	162.3	ug/L	722	Standard
	U	238	53.7	17.2	**0.0047**	0.001	26.0	ug/L	8	Standard
>	**Bi**	209	481410.0	0.8				ug/L	444563	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 11:31:14
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	35.0	62.3	**0.0064**	0.025	389.8	mg/L	38	Standard
K	39	21.7	35.3	**-0.0431**	0.030	69.7	mg/L	33	Standard
Ca	43	46.7	16.4	**0.7976**	1.093	137.1	mg/L	48	Standard
Fe	54	162.7	14.3	**0.0027**	0.018	653.5	mg/L	131	Standard
Fe	57	175.0	12.5	**0.0107**	0.058	546.2	mg/L	153	Standard
Sc-1	45	29386.8	0.7				mg/L	26715	Standard
Cl	35	118618.5	0.7				ug/L	102686	Standard
Kr	83	1.7	34.6				ug/L	2	Standard
Br	81	1786.8	0.6				ug/L	1897	Standard
P	31	20450.5	2.2				ug/L	18491	Standard
S	34	4707.4	8.1				ug/L	3734	Standard
Sr	88	51.7	14.8				ug/L	47	Standard
C	12	60.0	44.1				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	26.7	57.3				mg/L	23	Standard
Dy	164	35.4	117.1				mg/L	23	Standard
Ho-1	165	11.7	24.7				mg/L	7	Standard
Er	166	26.7	86.6				mg/L	13	Standard
I	127	10532.0	3.8				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		110.022	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 11:31:14
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.247
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.288
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 11:31:14
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606144511
Sample Date/Time: Tuesday, July 05, 2016 11:32:10
Number of Replicates: 3
Autosampler Position: 225
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	71021.2	3.7				ug/L	59517	Standard
	Be	9	871.7	8.6	**0.8097**	0.079	9.7	ug/L	10	Standard
	Al	27	12365222.8	2.8	**68.1084**	0.606	0.9	ug/L	1872	Standard
	Sc	45	34019.8	2.2				ug/L	26715	Standard
	Ti	47	2851.9	2.3	**10.3787**	0.373	3.6	ug/L	54	Standard
	V	51	136898.6	0.8	**16.6248**	0.202	1.2	ug/L	1419	Standard
	Cr	52	136908.1	0.6	**16.0816**	0.420	2.6	ug/L	7059	Standard
	Cr	53	18659.8	2.6	**18.1733**	0.240	1.3	ug/L	398	Standard
	Mn	55	3135205.7	0.4	**438.7022**	7.424	1.7	ug/L	824	Standard
	Co	59	69995.3	0.7	**9.0990**	0.118	1.3	ug/L	167	Standard
	Ni	60	42554.0	1.2	**19.9463**	0.236	1.2	ug/L	391	Standard
	Cu	65	26902.5	1.3	**12.6149**	0.241	1.9	ug/L	147	Standard
	Zn	66	77600.9	1.5	**62.1081**	0.974	1.6	ug/L	184	Standard
>	**Ge**	72	477668.6	2.0				ug/L	422998	Standard
	As	75	5139.4	0.7	**4.0643**	0.087	2.1	ug/L	-58	Standard
	Se	82	62.8	26.5	**0.2674**	0.120	44.8	ug/L	15	Standard
	Se-1	77	243.3	4.2	**2.1725**	0.065	3.0	ug/L	58	Standard
>	**Ga**	71	19127.1	0.7				mg/L	113	Standard
	Rb	85	198995.0	1.0				ug/L	27	Standard
	Y	89	649207.2	0.7				ug/L	383780	Standard
>	**Rh**	103	41.7	36.7				ug/L	18	Standard
	Mo	98	449.9	10.5	**0.1460**	0.016	10.7	ug/L	31	Standard
	Ag	107	632.0	3.3	**0.0739**	0.003	3.9	ug/L	84	Standard
	Cd	111	1171.9	1.1	**0.4997**	0.007	1.4	mg/L	4	Standard
	Cd	114	2756.6	4.8	**0.4838**	0.022	4.6	ug/L	15	Standard
>	**In**	115	562900.2	0.4				ug/L	515400	Standard
	Sn	118	1318.4	30.9	**0.0274**	0.064	232.7	ug/L	1043	Standard
	Sb	123	667.8	32.9	**0.1172**	0.044	37.3	ug/L	227	Standard
	Ba	135	585914.6	0.3	**266.8564**	1.879	0.7	ug/L	31	Standard
	Ce	140	1276381.6	0.5				ug/L	120	Standard
>	**Tb**	159	888751.2	0.9				ug/L	785817	Standard
	Ho	165	19744.5	3.6				ug/L	7	Standard
	Tl	203	1090.7	3.8	**0.1214**	0.004	3.5	ug/L	6	Standard
	Tl	205	1000.0	10.3	**0.1248**	0.014	10.9	ug/L	8	Standard
	Pb	206	500976.8	0.6	**92.3093**	1.215	1.3	ug/L	201	Standard
	Pb	207	419232.3	0.8	**85.4378**	0.957	1.1	ug/L	177	Standard
	Pb	208	1756286.6	0.3	**89.1076**	0.893	1.0	ug/L	722	Standard
	U	238	4139.9	1.2	**0.5503**	0.004	0.7	ug/L	8	Standard
>	**Bi**	209	484824.6	0.8				ug/L	444563	Standard

Sample ID: L1606144511
Report Date/Time: Tuesday, July 05, 2016 11:34:27
Page 1

Approved: July 06, 2016

Na	23	6.7	114.6	**60.6851**	68.539	112.9	mg/L	0	Standard
Mg	24	133.3	31.2	**0.0981**	0.042	42.4	mg/L	38	Standard
K	39	321.7	9.1	**0.9771**	0.107	10.9	mg/L	33	Standard
Ca	43	76.7	16.4	**3.5760**	1.345	37.6	mg/L	48	Standard
Fe	54	18043.1	2.5	**11.3093**	0.428	3.8	mg/L	131	Standard
Fe	57	4782.4	4.7	**11.1647**	0.761	6.8	mg/L	153	Standard
> **Sc-1**	45	34019.8	2.2				mg/L	26715	Standard
Cl	35	111550.6	1.9				ug/L	102686	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	3117.0	6.2				ug/L	1897	Standard
P	31	19804.6	0.2				ug/L	18491	Standard
S	34	4103.9	4.3				ug/L	3734	Standard
Sr	88	130.0	6.7				ug/L	47	Standard
C	12	120.0	16.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	60.0	28.9				mg/L	23	Standard
Dy	164	33281.5	0.5				mg/L	23	Standard
Ho-1	165	19744.5	3.6				mg/L	7	Standard
Er	166	18307.7	1.1				mg/L	13	Standard
I	127	22428.3	4.6				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		119.330	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		112.925	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144511
Report Date/Time: Tuesday, July 05, 2016 11:34:27
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		109.216
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		109.057
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606144511
Report Date/Time: Tuesday, July 05, 2016 11:34:27
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606144512
Sample Date/Time: Tuesday, July 05, 2016 11:35:21
Number of Replicates: 3
Autosampler Position: 226
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	71958.9	2.4				ug/L	59517	Standard
	Be	9	838.4	13.6	0.7678	0.123	16.0	ug/L	10	Standard
	Al	27	17699628.7	1.5	96.2224	1.478	1.5	ug/L	1872	Standard
	Sc	45	34681.3	2.5				ug/L	26715	Standard
	Ti	47	2998.3	2.5	10.8517	0.290	2.7	ug/L	54	Standard
	V	51	152944.2	1.3	18.4758	0.186	1.0	ug/L	1419	Standard
	Cr	52	107693.1	0.8	12.3552	0.145	1.2	ug/L	7059	Standard
	Cr	53	15069.2	1.0	14.5056	0.122	0.8	ug/L	398	Standard
	Mn	55	2049605.2	0.5	284.9255	0.634	0.2	ug/L	824	Standard
	Co	59	61986.5	0.5	8.0040	0.033	0.4	ug/L	167	Standard
	Ni	60	40334.6	0.6	18.7759	0.110	0.6	ug/L	391	Standard
	Cu	65	23405.4	0.8	10.8951	0.130	1.2	ug/L	147	Standard
	Zn	66	41678.2	0.7	33.0739	0.237	0.7	ug/L	184	Standard
>	Ge	72	480619.9	0.4				ug/L	422998	Standard
	As	75	4243.8	1.5	3.3439	0.060	1.8	ug/L	-58	Standard
	Se	82	52.0	15.0	0.1844	0.060	32.5	ug/L	15	Standard
	Se-1	77	217.0	7.2	1.8498	0.173	9.3	ug/L	58	Standard
>	Ga	71	16040.2	1.7				mg/L	113	Standard
	Rb	85	153835.4	1.2				ug/L	27	Standard
	Y	89	653034.0	1.2				ug/L	383780	Standard
>	Rh	103	31.7	9.1				ug/L	18	Standard
	Mo	98	421.3	2.5	0.1368	0.005	3.9	ug/L	31	Standard
	Ag	107	442.0	6.1	0.0483	0.003	6.0	ug/L	84	Standard
	Cd	111	237.2	8.6	0.0982	0.008	7.9	mg/L	4	Standard
	Cd	114	490.5	1.5	0.0827	0.002	2.6	ug/L	15	Standard
>	In	115	560411.2	1.6				ug/L	515400	Standard
	Sn	118	975.0	33.7	-0.0256	0.054	212.3	ug/L	1043	Standard
	Sb	123	548.8	20.4	0.0941	0.024	25.1	ug/L	227	Standard
	Ba	135	225197.1	1.5	103.0102	1.156	1.1	ug/L	31	Standard
	Ce	140	1202275.9	0.9				ug/L	120	Standard
>	Tb	159	880063.7	3.4				ug/L	785817	Standard
	Ho	165	19287.3	1.8				ug/L	7	Standard
	Tl	203	813.0	1.7	0.0888	0.002	2.7	ug/L	6	Standard
	Tl	205	713.4	9.0	0.0863	0.009	9.9	ug/L	8	Standard
	Pb	206	102726.4	1.2	18.7176	0.074	0.4	ug/L	201	Standard
	Pb	207	86210.6	1.8	17.3695	0.202	1.2	ug/L	177	Standard
	Pb	208	361707.3	1.0	18.1456	0.252	1.4	ug/L	722	Standard
	U	238	4018.9	1.5	0.5293	0.011	2.0	ug/L	8	Standard
>	Bi	209	489289.8	0.8				ug/L	444563	Standard

Sample ID: L1606144512
Report Date/Time: Tuesday, July 05, 2016 11:37:38
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**14.6402**	25.349	173.1	mg/L	0	Standard
Mg	24	108.3	14.1	**0.0713**	0.017	23.1	mg/L	38	Standard
K	39	311.7	16.1	**0.9211**	0.163	17.7	mg/L	33	Standard
Ca	43	93.3	17.2	**5.4689**	2.175	39.8	mg/L	48	Standard
Fe	54	14438.9	2.6	**8.8562**	0.446	5.0	mg/L	131	Standard
Fe	57	3728.8	4.6	**8.4284**	0.631	7.5	mg/L	153	Standard
Sc-1	45	34681.3	2.5				mg/L	26715	Standard
Cl	35	113693.7	1.3				ug/L	102686	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	2510.2	3.5				ug/L	1897	Standard
P	31	19771.2	2.0				ug/L	18491	Standard
S	34	4267.3	5.1				ug/L	3734	Standard
Sr	88	83.3	61.6				ug/L	47	Standard
C	12	126.7	45.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	43.3	48.0				mg/L	23	Standard
Dy	164	32527.2	3.4				mg/L	23	Standard
Ho-1	165	19287.3	1.8				mg/L	7	Standard
Er	166	17176.4	4.0				mg/L	13	Standard
I	127	11282.5	1.3				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		120.905	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		113.622	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606144512
Report Date/Time: Tuesday, July 05, 2016 11:37:38
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.733
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		110.061
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mn 55 Upper, S, EEE	Mn	55	
Ba 135 Upper, S, EEE	Ba	135	

Sample ID: L1606144512
Report Date/Time: Tuesday, July 05, 2016 11:37:38
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145101
Sample Date/Time: Tuesday, July 05, 2016 11:38:33
Number of Replicates: 3
Autosampler Position: 227
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	63086.2	3.9				ug/L	59517	Standard
	Be	9	48.3	58.8	**0.0207**	0.028	137.5	ug/L	10	Standard
	Al	27	187174.4	2.2	**1.1487**	0.023	2.0	ug/L	1872	Standard
	Sc	45	27850.6	4.7				ug/L	26715	Standard
	Ti	47	1001.4	3.2	**3.7116**	0.114	3.1	ug/L	54	Standard
	V	51	8090.4	4.8	**0.8746**	0.054	6.2	ug/L	1419	Standard
	Cr	52	11362.3	1.1	**0.5262**	0.019	3.6	ug/L	7059	Standard
	Cr	53	3108.7	1.2	**2.8761**	0.047	1.6	ug/L	398	Standard
	Mn	55	100253.9	1.0	**14.7608**	0.214	1.5	ug/L	824	Standard
	Co	59	1542.4	2.3	**0.1883**	0.004	2.2	ug/L	167	Standard
	Ni	60	1461.7	3.0	**0.5345**	0.020	3.7	ug/L	391	Standard
	Cu	65	1964.5	1.4	**0.9047**	0.018	2.0	ug/L	147	Standard
	Zn	66	8628.5	2.2	**7.1884**	0.148	2.1	ug/L	184	Standard
>	Ge	72	450022.4	0.4				ug/L	422998	Standard
	As	75	1048.0	2.6	**0.9200**	0.025	2.7	ug/L	-58	Standard
	Se	82	19.3	39.7	**-0.0513**	0.062	119.9	ug/L	15	Standard
	Se-1	77	150.3	2.5	**1.1949**	0.040	3.4	ug/L	58	Standard
>	Ga	71	646.7	5.1				mg/L	113	Standard
	Rb	85	4570.7	3.6				ug/L	27	Standard
	Y	89	391279.7	0.8				ug/L	383780	Standard
>	Rh	103	1.7	173.2				ug/L	18	Standard
	Mo	98	193.8	15.4	**0.0628**	0.012	18.9	ug/L	31	Standard
	Ag	107	106.3	6.0	**0.0035**	0.001	30.1	ug/L	84	Standard
	Cd	111	83.3	6.9	**0.0343**	0.002	6.3	mg/L	4	Standard
	Cd	114	206.4	14.3	**0.0347**	0.006	17.6	ug/L	15	Standard
>	In	115	524420.4	1.4				ug/L	515400	Standard
	Sn	118	508.3	21.2	**-0.0945**	0.019	20.6	ug/L	1043	Standard
	Sb	123	392.1	21.6	**0.0680**	0.019	28.3	ug/L	227	Standard
	Ba	135	9400.9	1.4	**4.5731**	0.093	2.0	ug/L	31	Standard
	Ce	140	55810.9	1.2				ug/L	120	Standard
>	Tb	159	818644.4	1.6				ug/L	785817	Standard
	Ho	165	396.7	1.5				ug/L	7	Standard
	Tl	203	74.0	20.2	**0.0052**	0.002	32.8	ug/L	6	Standard
	Tl	205	78.3	13.3	**0.0039**	0.001	34.4	ug/L	8	Standard
	Pb	206	23804.4	0.6	**4.5052**	0.016	0.3	ug/L	201	Standard
	Pb	207	19941.1	0.2	**4.1691**	0.037	0.9	ug/L	177	Standard
	Pb	208	83297.6	0.7	**4.3367**	0.052	1.2	ug/L	722	Standard
	U	238	385.3	3.7	**0.0509**	0.002	4.1	ug/L	8	Standard
>	Bi	209	467495.4	0.9				ug/L	444563	Standard

Sample ID: L1606145101
Report Date/Time: Tuesday, July 05, 2016 11:40:49
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	41.7	56.7	**0.0172**	0.029	170.4	mg/L	38	Standard
K	39	41.7	30.2	**0.0444**	0.045	101.3	mg/L	33	Standard
Ca	43	20.0	50.0	**-2.8724**	1.432	49.8	mg/L	48	Standard
Fe	54	717.3	6.4	**0.4381**	0.026	5.9	mg/L	131	Standard
Fe	57	350.0	22.0	**0.5664**	0.274	48.3	mg/L	153	Standard
> **Sc-1**	45	27850.6	4.7				mg/L	26715	Standard
Cl	35	101597.5	0.2				ug/L	102686	Standard
Kr	83	3.7	56.8				ug/L	2	Standard
Br	81	1566.7	0.7				ug/L	1897	Standard
P	31	7351.8	5.4				ug/L	18491	Standard
S	34	3560.4	3.1				ug/L	3734	Standard
Sr	88	83.3	54.1				ug/L	47	Standard
C	12	73.3	15.7				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	30.0	33.3				mg/L	23	Standard
Dy	164	819.3	12.4				mg/L	23	Standard
Ho-1	165	396.7	1.5				mg/L	7	Standard
Er	166	433.3	1.3				mg/L	13	Standard
I	127	25947.2	1.1				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		105.997	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		106.389	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145101
Report Date/Time: Tuesday, July 05, 2016 11:40:49
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.750
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.158
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145102
Sample Date/Time: Tuesday, July 05, 2016 11:41:44
Number of Replicates: 3
Autosampler Position: 228
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	61941.5	4.6				ug/L	59517	Standard
	Be	9	58.3	9.9	**0.0335**	0.008	23.1	ug/L	10	Standard
	Al	27	397101.0	1.9	**2.4976**	0.075	3.0	ug/L	1872	Standard
	Sc	45	27464.9	1.3				ug/L	26715	Standard
	Ti	47	1832.8	2.5	**7.0385**	0.120	1.7	ug/L	54	Standard
	V	51	14851.8	2.2	**1.7665**	0.058	3.3	ug/L	1419	Standard
	Cr	52	16937.1	0.9	**1.2760**	0.042	3.3	ug/L	7059	Standard
	Cr	53	3678.8	5.9	**3.5022**	0.267	7.6	ug/L	398	Standard
	Mn	55	333369.6	0.1	**49.6668**	0.486	1.0	ug/L	824	Standard
	Co	59	4435.3	0.3	**0.5920**	0.006	1.0	ug/L	167	Standard
	Ni	60	2891.6	2.6	**1.2612**	0.037	2.9	ug/L	391	Standard
	Cu	65	5814.1	0.3	**2.8489**	0.035	1.2	ug/L	147	Standard
	Zn	66	20833.6	1.1	**17.6837**	0.331	1.9	ug/L	184	Standard
>	Ge	72	447514.4	1.0				ug/L	422998	Standard
	As	75	928.7	4.0	**0.8257**	0.039	4.7	ug/L	-58	Standard
	Se	82	23.8	9.6	**-0.0141**	0.020	145.1	ug/L	15	Standard
	Se-1	77	153.3	7.5	**1.2422**	0.127	10.2	ug/L	58	Standard
>	Ga	71	1306.7	7.7				mg/L	113	Standard
	Rb	85	13222.4	2.0				ug/L	27	Standard
	Y	89	388639.7	0.5				ug/L	383780	Standard
>	Rh	103	13.3	78.1				ug/L	18	Standard
	Mo	98	448.8	3.5	**0.1576**	0.004	2.6	ug/L	31	Standard
	Ag	107	124.3	5.2	**0.0062**	0.001	18.2	ug/L	84	Standard
	Cd	111	98.1	12.8	**0.0413**	0.006	13.6	mg/L	4	Standard
	Cd	114	206.1	2.6	**0.0348**	0.001	4.0	ug/L	15	Standard
>	In	115	522465.9	1.3				ug/L	515400	Standard
	Sn	118	518.3	40.3	**-0.0928**	0.035	37.3	ug/L	1043	Standard
	Sb	123	307.8	24.2	**0.0500**	0.016	31.6	ug/L	227	Standard
	Ba	135	21685.8	0.0	**10.6194**	0.135	1.3	ug/L	31	Standard
	Ce	140	63251.8	1.4				ug/L	120	Standard
>	Tb	159	805776.3	2.0				ug/L	785817	Standard
	Ho	165	756.7	9.4				ug/L	7	Standard
	Tl	203	171.7	5.0	**0.0170**	0.001	7.0	ug/L	6	Standard
	Tl	205	131.7	12.2	**0.0113**	0.002	20.2	ug/L	8	Standard
	Pb	206	15597.4	1.2	**2.9590**	0.060	2.0	ug/L	201	Standard
	Pb	207	12686.3	1.2	**2.6552**	0.046	1.7	ug/L	177	Standard
	Pb	208	54000.2	0.6	**2.8156**	0.019	0.7	ug/L	722	Standard
	U	238	900.4	1.8	**0.1232**	0.003	2.7	ug/L	8	Standard
>	Bi	209	463997.6	0.9				ug/L	444563	Standard

Sample ID: L1606145102
Report Date/Time: Tuesday, July 05, 2016 11:44:01
Page 1

Approved: July 06, 2016

Na	23	3.3	86.6	**37.9829**	32.899	86.6	mg/L	0	Standard
Mg	24	41.7	36.7	**0.0174**	0.019	110.5	mg/L	38	Standard
K	39	38.3	61.6	**0.0342**	0.101	294.9	mg/L	33	Standard
Ca	43	43.3	35.3	**0.7488**	2.326	310.6	mg/L	48	Standard
Fe	54	1680.0	5.2	**1.2006**	0.065	5.4	mg/L	131	Standard
Fe	57	590.0	9.0	**1.2966**	0.165	12.8	mg/L	153	Standard
> **Sc-1**	45	27464.9	1.3				mg/L	26715	Standard
Cl	35	103666.1	1.2				ug/L	102686	Standard
Kr	83	3.3	45.8				ug/L	2	Standard
Br	81	1920.1	6.8				ug/L	1897	Standard
P	31	7571.9	4.4				ug/L	18491	Standard
S	34	3707.1	3.8				ug/L	3734	Standard
Sr	88	76.7	43.4				ug/L	47	Standard
C	12	83.3	61.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	16.7	34.6				mg/L	23	Standard
Dy	164	1215.4	9.0				mg/L	23	Standard
Ho-1	165	756.7	9.4				mg/L	7	Standard
Er	166	723.4	7.6				mg/L	13	Standard
I	127	26428.0	1.1				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		104.074	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		105.796	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145102
Report Date/Time: Tuesday, July 05, 2016 11:44:01
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.371
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.372
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145102PS WG575057-01
Sample Date/Time: Tuesday, July 05, 2016 11:44:55
Number of Replicates: 3
Autosampler Position: 229
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	62697.9	4.2				ug/L	59517	Standard
	Be	9	44294.7	3.5	**48.3430**	1.008	2.1	ug/L	10	Standard
	Al	27	380968.8	3.5	**2.3649**	0.024	1.0	ug/L	1872	Standard
	Sc	45	27576.7	2.5				ug/L	26715	Standard
	Ti	47	1757.8	3.6	**6.6590**	0.216	3.3	ug/L	54	Standard
	V	51	369561.9	0.1	**47.6704**	0.227	0.5	ug/L	1419	Standard
	Cr	52	375004.1	0.6	**48.2990**	0.392	0.8	ug/L	7059	Standard
	Cr	53	47950.8	1.0	**49.9721**	0.632	1.3	ug/L	398	Standard
	Mn	55	624219.2	1.1	**92.0208**	0.683	0.7	ug/L	824	Standard
	Co	59	331225.4	0.7	**45.5157**	0.136	0.3	ug/L	167	Standard
	Ni	60	98193.5	1.1	**48.8356**	0.402	0.8	ug/L	391	Standard
	Cu	65	103660.4	1.5	**51.5103**	0.662	1.3	ug/L	147	Standard
	Zn	66	82546.9	0.1	**69.7085**	0.295	0.4	ug/L	184	Standard
>	Ge	72	452783.9	0.5				ug/L	422998	Standard
	As	75	61959.6	0.0	**51.0727**	0.255	0.5	ug/L	-58	Standard
	Se	82	6718.1	0.6	**53.3265**	0.486	0.9	ug/L	15	Standard
	Se-1	77	4447.3	1.5	**54.1235**	1.057	2.0	ug/L	58	Standard
>	Ga	71	1278.4	1.3				mg/L	113	Standard
	Rb	85	12908.8	1.4				ug/L	27	Standard
	Y	89	401631.9	0.4				ug/L	383780	Standard
>	Rh	103	31.7	50.8				ug/L	18	Standard
	Mo	98	450.5	1.2	**0.1549**	0.001	1.0	ug/L	31	Standard
	Ag	107	331680.1	0.3	**47.5704**	0.222	0.5	ug/L	84	Standard
	Cd	111	108869.6	0.6	**49.4140**	0.475	1.0	mg/L	4	Standard
	Cd	114	263950.9	0.6	**49.3701**	0.312	0.6	ug/L	15	Standard
>	In	115	533128.8	0.5				ug/L	515400	Standard
	Sn	118	636.7	25.1	**-0.0747**	0.026	35.4	ug/L	1043	Standard
	Sb	123	196248.0	6.3	**41.4359**	2.744	6.6	ug/L	227	Standard
	Ba	135	117350.4	0.6	**56.4148**	0.661	1.2	ug/L	31	Standard
	Ce	140	61538.1	0.5				ug/L	120	Standard
>	Tb	159	822948.5	0.5				ug/L	785817	Standard
	Ho	165	756.7	6.0				ug/L	7	Standard
	Tl	203	396729.2	0.3	**46.1565**	0.208	0.4	ug/L	6	Standard
	Tl	205	348011.4	0.7	**46.4767**	0.241	0.5	ug/L	8	Standard
	Pb	206	267711.4	0.3	**50.0339**	0.138	0.3	ug/L	201	Standard
	Pb	207	232867.9	0.4	**48.1364**	0.017	0.0	ug/L	177	Standard
	Pb	208	960047.3	0.6	**49.4056**	0.263	0.5	ug/L	722	Standard
	U	238	326850.5	0.2	**44.2871**	0.244	0.6	ug/L	8	Standard
>	Bi	209	477758.5	0.4				ug/L	444563	Standard

Sample ID: L1606145102PS WG575057-01
Report Date/Time: Tuesday, July 05, 2016 11:47:12
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**18.9251**	32.771	173.2	mg/L	0	Standard
Mg	24	38.3	27.2	**0.0131**	0.013	98.6	mg/L	38	Standard
K	39	41.7	18.3	**0.0472**	0.029	62.1	mg/L	33	Standard
Ca	43	43.3	40.5	**0.7741**	2.837	366.4	mg/L	48	Standard
Fe	54	1650.4	2.6	**1.1722**	0.005	0.4	mg/L	131	Standard
Fe	57	556.7	10.4	**1.1920**	0.207	17.4	mg/L	153	Standard
> **Sc-1**	45	27576.7	2.5				mg/L	26715	Standard
Cl	35	103188.9	0.2				ug/L	102686	Standard
Kr	83	2.3	89.2				ug/L	2	Standard
Br	81	1846.8	10.9				ug/L	1897	Standard
P	31	7151.7	0.7				ug/L	18491	Standard
S	34	3632.1	3.8				ug/L	3734	Standard
Sr	88	60.0	30.0				ug/L	47	Standard
C	12	90.0	66.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	23.3	49.5				mg/L	23	Standard
Dy	164	1235.3	9.5				mg/L	23	Standard
Ho-1	165	756.7	6.0				mg/L	7	Standard
Er	166	656.7	24.8				mg/L	13	Standard
I	127	26660.1	0.7				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		105.345	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		107.042	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145102PS WG575057-01
Report Date/Time: Tuesday, July 05, 2016 11:47:12
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		103.440
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		107.467
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145102SDL WG575057-02
Sample Date/Time: Tuesday, July 05, 2016 11:48:07
Number of Replicates: 3
Autosampler Position: 230
Sample Description: ~~10~~ 50
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	54039.6	3.1				ug/L	59517	Standard
	Be	9	48.3	58.8	**0.0295**	0.034	115.5	ug/L	10	Standard
	Al	27	89003.2	15.7	**0.6307**	0.086	13.6	ug/L	1872	Standard
	Sc	45	23686.8	0.3				ug/L	26715	Standard
	Ti	47	378.0	5.6	**1.4703**	0.090	6.2	ug/L	54	Standard
	V	51	4224.3	4.5	**0.4528**	0.023	5.1	ug/L	1419	Standard
	Cr	52	7865.0	4.3	**0.2187**	0.042	19.2	ug/L	7059	Standard
	Cr	53	2401.9	6.7	**2.5065**	0.166	6.6	ug/L	398	Standard
	Mn	55	67918.0	3.7	**11.4603**	0.319	2.8	ug/L	824	Standard
	Co	59	1167.7	13.0	**0.1604**	0.022	13.8	ug/L	167	Standard
	Ni	60	836.0	11.2	**0.2826**	0.050	17.8	ug/L	391	Standard
	Cu	65	1282.4	2.5	**0.6589**	0.017	2.6	ug/L	147	Standard
	Zn	66	5007.5	2.9	**4.7407**	0.108	2.3	ug/L	184	Standard
>	**Ge**	72	391606.9	1.0				ug/L	422998	Standard
	As	75	174.6	23.0	**0.2178**	0.037	17.2	ug/L	-58	Standard
	Se	82	20.5	32.2	**-0.0173**	0.060	344.9	ug/L	15	Standard
	Se-1	77	133.3	1.9	**1.2312**	0.056	4.5	ug/L	58	Standard
>	**Ga**	71	268.3	15.1				mg/L	113	Standard
	Rb	85	2626.9	3.6				ug/L	27	Standard
	Y	89	337426.2	1.1				ug/L	383780	Standard
>	**Rh**	103	15.0	66.7				ug/L	18	Standard
	Mo	98	156.5	9.3	**0.0577**	0.006	10.6	ug/L	31	Standard
	Ag	107	191.0	71.7	**0.0200**	0.023	113.4	ug/L	84	Standard
	Cd	111	50.3	69.4	**0.0225**	0.018	81.1	mg/L	4	Standard
	Cd	114	171.3	36.7	**0.0329**	0.013	40.7	ug/L	15	Standard
>	**In**	115	455091.8	1.2				ug/L	515400	Standard
	Sn	118	385.0	19.7	**-0.1055**	0.016	14.9	ug/L	1043	Standard
	Sb	123	2217.5	20.0	**0.5330**	0.115	21.6	ug/L	227	Standard
	Ba	135	4570.0	8.0	**2.5499**	0.181	7.1	ug/L	31	Standard
	Ce	140	13090.7	6.8				ug/L	120	Standard
>	**Tb**	159	708958.3	1.6				ug/L	785817	Standard
	Ho	165	178.3	12.6				ug/L	7	Standard
	Tl	203	212.3	44.0	**0.0250**	0.012	49.8	ug/L	6	Standard
	Tl	205	268.3	50.7	**0.0348**	0.021	60.0	ug/L	8	Standard
	Pb	206	3361.7	9.7	**0.6843**	0.068	10.0	ug/L	201	Standard
	Pb	207	2820.3	6.6	**0.6295**	0.043	6.8	ug/L	177	Standard
	Pb	208	12064.2	11.5	**0.6727**	0.080	12.0	ug/L	722	Standard
	U	238	331.0	42.4	**0.0495**	0.022	44.0	ug/L	8	Standard
>	**Bi**	209	411532.0	0.4				ug/L	444563	Standard

Sample ID: L1606145102SDL WG575057-02
Report Date/Time: Tuesday, July 05, 2016 11:50:23
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**21.9603**	38.028	173.2	mg/L	0	Standard
Mg	24	35.0	65.5	**0.0159**	0.032	203.8	mg/L	38	Standard
K	39	50.0	40.0	**0.1177**	0.099	84.4	mg/L	33	Standard
Ca	43	31.7	50.8	**-0.2632**	2.835	1077.0	mg/L	48	Standard
Fe	54	367.4	21.7	**0.2172**	0.072	32.9	mg/L	131	Standard
Fe	57	280.0	18.8	**0.4969**	0.187	37.7	mg/L	153	Standard
> **Sc-1**	45	23686.8	0.3				mg/L	26715	Standard
Cl	35	99858.7	1.1				ug/L	102686	Standard
Kr	83	1.3	86.6				ug/L	2	Standard
Br	81	1640.1	12.5				ug/L	1897	Standard
P	31	5642.7	5.7				ug/L	18491	Standard
S	34	3278.7	2.2				ug/L	3734	Standard
Sr	88	75.0	11.5				ug/L	47	Standard
C	12	36.7	15.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	23	Standard
Dy	164	282.5	17.5				mg/L	23	Standard
Ho-1	165	178.3	12.6				mg/L	7	Standard
Er	166	156.7	35.2				mg/L	13	Standard
I	127	28024.2	2.4				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		90.797	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		92.579	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145102SDL WG575057-02
Report Date/Time: Tuesday, July 05, 2016 11:50:23
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		88.299
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		92.570
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: L1606145102SDL WG575057-02
Report Date/Time: Tuesday, July 05, 2016 11:50:23
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145103
Sample Date/Time: Tuesday, July 05, 2016 11:51:18
Number of Replicates: 3
Autosampler Position: 231
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	58994.6	1.3				ug/L	59517	Standard
	Be	9	38.3	37.7	**0.0135**	0.017	129.1	ug/L	10	Standard
	Al	27	155374.4	1.4	**1.0179**	0.018	1.7	ug/L	1872	Standard
	Sc	45	25635.0	4.0				ug/L	26715	Standard
	Ti	47	1269.4	0.7	**5.1547**	0.113	2.2	ug/L	54	Standard
	V	51	9899.5	0.9	**1.2086**	0.027	2.2	ug/L	1419	Standard
	Cr	52	12883.1	0.5	**0.8572**	0.024	2.9	ug/L	7059	Standard
	Cr	53	2881.9	5.1	**2.8672**	0.116	4.0	ug/L	398	Standard
	Mn	55	158042.5	3.0	**25.1279**	0.371	1.5	ug/L	824	Standard
	Co	59	1946.5	3.1	**0.2647**	0.005	2.0	ug/L	167	Standard
	Ni	60	1707.1	5.4	**0.7229**	0.038	5.3	ug/L	391	Standard
	Cu	65	2308.8	1.9	**1.1652**	0.014	1.2	ug/L	147	Standard
	Zn	66	13053.3	0.7	**11.8043**	0.102	0.9	ug/L	184	Standard
>	**Ge**	72	418173.0	1.6				ug/L	422998	Standard
	As	75	531.5	13.1	**0.5253**	0.059	11.1	ug/L	-58	Standard
	Se	82	22.2	54.0	**-0.0153**	0.101	659.9	ug/L	15	Standard
	Se-1	77	136.3	5.9	**1.1512**	0.127	11.0	ug/L	58	Standard
>	**Ga**	71	790.0	11.2				mg/L	113	Standard
	Rb	85	6226.3	6.4				ug/L	27	Standard
	Y	89	366669.1	1.2				ug/L	383780	Standard
>	**Rh**	103	28.3	36.7				ug/L	18	Standard
	Mo	98	322.8	6.1	**0.1172**	0.008	6.8	ug/L	31	Standard
	Ag	107	106.3	3.9	**0.0044**	0.001	11.5	ug/L	84	Standard
	Cd	111	101.0	5.1	**0.0452**	0.002	4.9	mg/L	4	Standard
	Cd	114	217.0	6.1	**0.0390**	0.002	5.6	ug/L	15	Standard
>	**In**	115	496031.9	1.1				ug/L	515400	Standard
	Sn	118	485.0	36.9	**-0.0941**	0.031	33.0	ug/L	1043	Standard
	Sb	123	1006.7	28.3	**0.2117**	0.062	29.3	ug/L	227	Standard
	Ba	135	10215.4	2.6	**5.2559**	0.079	1.5	ug/L	31	Standard
	Ce	140	49659.6	2.7				ug/L	120	Standard
>	**Tb**	159	759583.5	0.8				ug/L	785817	Standard
	Ho	165	383.3	14.1				ug/L	7	Standard
	Tl	203	121.0	13.0	**0.0117**	0.002	16.0	ug/L	6	Standard
	Tl	205	131.7	7.9	**0.0123**	0.002	12.8	ug/L	8	Standard
	Pb	206	31584.6	1.8	**6.3716**	0.110	1.7	ug/L	201	Standard
	Pb	207	26494.8	0.6	**5.9067**	0.071	1.2	ug/L	177	Standard
	Pb	208	110332.0	2.1	**6.1238**	0.088	1.4	ug/L	722	Standard
	U	238	469.3	1.2	**0.0666**	0.001	1.3	ug/L	8	Standard
>	**Bi**	209	439883.7	0.8				ug/L	444563	Standard

Sample ID: L1606145103
Report Date/Time: Tuesday, July 05, 2016 11:53:35
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	30.0	72.6	**0.0064**	0.030	478.1	mg/L	38	Standard
K	39	40.0	37.5	**0.0525**	0.064	121.8	mg/L	33	Standard
Ca	43	33.3	34.6	**-0.3638**	2.049	563.3	mg/L	48	Standard
Fe	54	1092.3	3.6	**0.8011**	0.010	1.2	mg/L	131	Standard
Fe	57	438.3	6.9	**0.9374**	0.154	16.4	mg/L	153	Standard
> **Sc-1**	45	25635.0	4.0				mg/L	26715	Standard
Cl	35	101103.6	1.2				ug/L	102686	Standard
Kr	83	1.7	91.7				ug/L	2	Standard
Br	81	1673.4	12.5				ug/L	1897	Standard
P	31	7090.0	4.6				ug/L	18491	Standard
S	34	3425.4	0.4				ug/L	3734	Standard
Sr	88	58.3	27.6				ug/L	47	Standard
C	12	70.0	62.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	100.0				mg/L	23	Standard
Dy	164	602.6	18.5				mg/L	23	Standard
Ho-1	165	383.3	14.1				mg/L	7	Standard
Er	166	363.3	8.4				mg/L	13	Standard
I	127	30303.6	1.9				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		99.123	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		98.859	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145103
Report Date/Time: Tuesday, July 05, 2016 11:53:35
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.242
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.947
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145104
Sample Date/Time: Tuesday, July 05, 2016 11:54:30
Number of Replicates: 3
Autosampler Position: 232
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	59640.5	2.3				ug/L	59517	Standard
	Be	9	51.7	14.8	**0.0282**	0.009	32.9	ug/L	10	Standard
	Al	27	651313.8	1.1	**4.2602**	0.050	1.2	ug/L	1872	Standard
	Sc	45	27301.3	2.9				ug/L	26715	Standard
	Ti	47	1316.1	3.5	**5.0645**	0.163	3.2	ug/L	54	Standard
	V	51	8314.8	1.0	**0.9265**	0.012	1.3	ug/L	1419	Standard
	Cr	52	12316.3	0.6	**0.6864**	0.002	0.3	ug/L	7059	Standard
	Cr	53	2606.9	2.8	**2.4032**	0.063	2.6	ug/L	398	Standard
	Mn	55	151787.5	0.5	**22.8881**	0.145	0.6	ug/L	824	Standard
	Co	59	2591.6	1.2	**0.3410**	0.005	1.4	ug/L	167	Standard
	Ni	60	1900.5	2.9	**0.7750**	0.023	3.0	ug/L	391	Standard
	Cu	65	2716.6	1.1	**1.3098**	0.008	0.6	ug/L	147	Standard
	Zn	66	12752.4	1.0	**10.9286**	0.144	1.3	ug/L	184	Standard
>	Ge	72	440774.5	0.5				ug/L	422998	Standard
	As	75	461.4	9.2	**0.4420**	0.038	8.6	ug/L	-58	Standard
	Se	82	21.9	26.2	**-0.0264**	0.048	180.8	ug/L	15	Standard
	Se-1	77	123.7	10.6	**0.8972**	0.173	19.2	ug/L	58	Standard
>	Ga	71	743.4	6.6				mg/L	113	Standard
	Rb	85	5377.6	4.8				ug/L	27	Standard
	Y	89	383304.5	0.0				ug/L	383780	Standard
>	Rh	103	15.0	57.7				ug/L	18	Standard
	Mo	98	416.2	6.8	**0.1460**	0.012	8.5	ug/L	31	Standard
	Ag	107	96.0	9.5	**0.0021**	0.001	69.4	ug/L	84	Standard
	Cd	111	76.9	16.8	**0.0316**	0.006	19.0	mg/L	4	Standard
	Cd	114	202.2	26.9	**0.0341**	0.010	30.5	ug/L	15	Standard
>	In	115	521154.6	1.3				ug/L	515400	Standard
	Sn	118	476.7	10.6	**-0.0995**	0.010	9.8	ug/L	1043	Standard
	Sb	123	713.8	30.5	**0.1382**	0.049	35.3	ug/L	227	Standard
	Ba	135	9897.2	1.6	**4.8465**	0.129	2.7	ug/L	31	Standard
	Ce	140	49096.1	2.3				ug/L	120	Standard
>	Tb	159	791948.4	0.3				ug/L	785817	Standard
	Ho	165	413.3	12.2				ug/L	7	Standard
	Tl	203	144.0	10.8	**0.0137**	0.002	14.1	ug/L	6	Standard
	Tl	205	81.7	40.8	**0.0045**	0.005	102.1	ug/L	8	Standard
	Pb	206	24670.1	1.4	**4.7209**	0.060	1.3	ug/L	201	Standard
	Pb	207	20384.7	1.5	**4.3087**	0.070	1.6	ug/L	177	Standard
	Pb	208	86971.8	0.8	**4.5787**	0.029	0.6	ug/L	722	Standard
	U	238	541.0	5.2	**0.0732**	0.004	5.7	ug/L	8	Standard
>	Bi	209	462557.2	0.3				ug/L	444563	Standard

Sample ID: L1606145104
Report Date/Time: Tuesday, July 05, 2016 11:56:46
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	28.3	97.2	**0.0015**	0.034	2280.0	mg/L	38	Standard
K	39	41.7	25.0	**0.0492**	0.045	91.0	mg/L	33	Standard
Ca	43	28.3	10.2	**-1.5191**	0.388	25.5	mg/L	48	Standard
Fe	54	952.3	2.4	**0.6352**	0.036	5.6	mg/L	131	Standard
Fe	57	405.0	3.3	**0.7465**	0.048	6.5	mg/L	153	Standard
Sc-1	45	27301.3	2.9				mg/L	26715	Standard
Cl	35	106909.4	2.8				ug/L	102686	Standard
Kr	83	2.7	78.1				ug/L	2	Standard
Br	81	1606.8	7.6				ug/L	1897	Standard
P	31	7366.8	2.0				ug/L	18491	Standard
S	34	3655.4	3.4				ug/L	3734	Standard
Sr	88	70.0	21.4				ug/L	47	Standard
C	12	70.0	37.8				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	13.3	43.3				mg/L	23	Standard
Dy	164	731.8	10.6				mg/L	23	Standard
Ho-1	165	413.3	12.2				mg/L	7	Standard
Er	166	450.0	12.4				mg/L	13	Standard
I	127	29051.2	0.4				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6		100.208	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		104.203	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: L1606145104
Report Date/Time: Tuesday, July 05, 2016 11:56:46
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		101.117
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		104.048
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145105
Sample Date/Time: Tuesday, July 05, 2016 11:57:41
Number of Replicates: 3
Autosampler Position: 233
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	58588.2	3.7				ug/L	59517	Standard
	Be	9	116.7	31.0	**0.1053**	0.042	40.2	ug/L	10	Standard
	Al	27	92826.9	1.9	**0.6081**	0.031	5.1	ug/L	1872	Standard
	Sc	45	26227.7	1.7				ug/L	26715	Standard
	Ti	47	3936.5	2.1	**16.3111**	0.457	2.8	ug/L	54	Standard
	V	51	10928.0	1.0	**1.3351**	0.013	1.0	ug/L	1419	Standard
	Cr	52	17058.9	1.2	**1.4233**	0.022	1.6	ug/L	7059	Standard
	Cr	53	3093.7	6.6	**3.0693**	0.235	7.6	ug/L	398	Standard
	Mn	55	27000.7	1.1	**4.1360**	0.069	1.7	ug/L	824	Standard
	Co	59	1487.4	1.9	**0.1938**	0.005	2.4	ug/L	167	Standard
	Ni	60	1703.4	3.2	**0.7106**	0.025	3.5	ug/L	391	Standard
	Cu	65	2105.5	2.7	**1.0428**	0.036	3.5	ug/L	147	Standard
	Zn	66	8804.9	1.4	**7.8184**	0.158	2.0	ug/L	184	Standard
>	Ge	72	423003.0	0.7				ug/L	422998	Standard
	As	75	1338.6	0.3	**1.2315**	0.007	0.6	ug/L	-58	Standard
	Se	82	21.3	5.1	**-0.0243**	0.011	43.4	ug/L	15	Standard
	Se-1	77	132.3	4.8	**1.0768**	0.084	7.8	ug/L	58	Standard
>	Ga	71	1020.0	4.5				mg/L	113	Standard
	Rb	85	5507.7	0.7				ug/L	27	Standard
	Y	89	373045.0	0.6				ug/L	383780	Standard
>	Rh	103	11.7	137.8				ug/L	18	Standard
	Mo	98	262.3	6.0	**0.0924**	0.006	6.9	ug/L	31	Standard
	Ag	107	109.7	1.4	**0.0047**	0.000	6.0	ug/L	84	Standard
	Cd	111	109.5	5.6	**0.0487**	0.003	5.7	mg/L	4	Standard
	Cd	114	282.5	31.4	**0.0516**	0.018	34.4	ug/L	15	Standard
>	In	115	501649.9	0.3				ug/L	515400	Standard
	Sn	118	526.7	18.7	**-0.0875**	0.018	20.2	ug/L	1043	Standard
	Sb	123	586.8	28.7	**0.1154**	0.038	32.8	ug/L	227	Standard
	Ba	135	9884.6	1.2	**5.0283**	0.069	1.4	ug/L	31	Standard
	Ce	140	26367.9	3.0				ug/L	120	Standard
>	Tb	159	777177.3	0.4				ug/L	785817	Standard
	Ho	165	535.0	9.5				ug/L	7	Standard
	Tl	203	451.3	3.8	**0.0526**	0.002	3.9	ug/L	6	Standard
	Tl	205	398.3	22.1	**0.0502**	0.012	24.8	ug/L	8	Standard
	Pb	206	14523.0	1.7	**2.8646**	0.053	1.9	ug/L	201	Standard
	Pb	207	12095.5	2.0	**2.6327**	0.025	0.9	ug/L	177	Standard
	Pb	208	50729.4	1.8	**2.7505**	0.048	1.8	ug/L	722	Standard
	U	238	404.3	2.6	**0.0562**	0.001	2.3	ug/L	8	Standard
>	Bi	209	446032.8	1.0				ug/L	444563	Standard

Sample ID: L1606145105
Report Date/Time: Tuesday, July 05, 2016 11:59:58
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	43.3	29.0	**0.0218**	0.016	73.7	mg/L	38	Standard
K	39	43.3	46.6	**0.0650**	0.094	144.6	mg/L	33	Standard
Ca	43	41.7	48.5	**0.7656**	3.153	411.9	mg/L	48	Standard
Fe	54	1227.0	6.0	**0.8915**	0.073	8.2	mg/L	131	Standard
Fe	57	510.0	2.9	**1.1274**	0.028	2.4	mg/L	153	Standard
> **Sc-1**	45	26227.7	1.7				mg/L	26715	Standard
Cl	35	106266.0	1.0				ug/L	102686	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	1563.4	5.4				ug/L	1897	Standard
P	31	7021.6	1.0				ug/L	18491	Standard
S	34	3633.8	1.6				ug/L	3734	Standard
Sr	88	80.0	16.5				ug/L	47	Standard
C	12	60.0	28.9				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	23	Standard
Dy	164	911.9	3.4				mg/L	23	Standard
Ho-1	165	535.0	9.5				mg/L	7	Standard
Er	166	516.7	15.8				mg/L	13	Standard
I	127	29779.2	1.7				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		98.440	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.001	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145105
Report Date/Time: Tuesday, July 05, 2016 11:59:58
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		97.332
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.331
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: L1606145105
Report Date/Time: Tuesday, July 05, 2016 11:59:58
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145106 WG574521-02
Sample Date/Time: Tuesday, July 05, 2016 12:00:52
Number of Replicates: 3
Autosampler Position: 234
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	58452.6	3.5				ug/L	59517	Standard
	Be	9	111.7	38.1	**0.1002**	0.053	52.4	ug/L	10	Standard
	Al	27	318919.9	0.7	**2.1238**	0.084	4.0	ug/L	1872	Standard
	Sc	45	25932.2	2.7				ug/L	26715	Standard
	Ti	47	2048.1	2.1	**8.3389**	0.301	3.6	ug/L	54	Standard
	V	51	42887.2	0.5	**5.7438**	0.096	1.7	ug/L	1419	Standard
	Cr	52	18270.0	1.2	**1.5847**	0.044	2.8	ug/L	7059	Standard
	Cr	53	3193.7	2.4	**3.1683**	0.035	1.1	ug/L	398	Standard
	Mn	55	163898.4	0.6	**25.6789**	0.411	1.6	ug/L	824	Standard
	Co	59	2641.9	0.7	**0.3624**	0.009	2.3	ug/L	167	Standard
	Ni	60	4286.3	1.6	**2.0837**	0.070	3.4	ug/L	391	Standard
	Cu	65	9810.8	2.0	**5.1286**	0.077	1.5	ug/L	147	Standard
	Zn	66	10912.9	1.2	**9.6933**	0.179	1.9	ug/L	184	Standard
>	Ge	72	424537.0	1.5				ug/L	422998	Standard
	As	75	1501.2	0.8	**1.3704**	0.029	2.1	ug/L	-58	Standard
	Se	82	59.7	8.9	**0.3007**	0.039	12.8	ug/L	15	Standard
	Se-1	77	144.7	3.1	**1.2324**	0.051	4.1	ug/L	58	Standard
>	Ga	71	1151.7	10.0				mg/L	113	Standard
	Rb	85	4672.4	3.9				ug/L	27	Standard
	Y	89	398677.4	1.0				ug/L	383780	Standard
>	Rh	103	15.0	57.7				ug/L	18	Standard
	Mo	98	790.9	3.3	**0.2985**	0.008	2.7	ug/L	31	Standard
	Ag	107	125.7	19.7	**0.0073**	0.004	55.8	ug/L	84	Standard
	Cd	111	462.9	4.3	**0.2204**	0.007	3.2	mg/L	4	Standard
	Cd	114	1087.2	1.9	**0.2129**	0.007	3.2	ug/L	15	Standard
>	In	115	498713.7	1.6				ug/L	515400	Standard
	Sn	118	418.3	38.7	**-0.1061**	0.029	27.6	ug/L	1043	Standard
	Sb	123	472.2	20.4	**0.0905**	0.023	25.5	ug/L	227	Standard
	Ba	135	25027.3	1.0	**12.8465**	0.328	2.6	ug/L	31	Standard
	Ce	140	130403.8	0.7				ug/L	120	Standard
>	Tb	159	771205.1	1.2				ug/L	785817	Standard
	Ho	165	3028.6	4.1				ug/L	7	Standard
	Tl	203	207.7	5.7	**0.0222**	0.002	8.0	ug/L	6	Standard
	Tl	205	175.0	8.6	**0.0181**	0.002	10.2	ug/L	8	Standard
	Pb	206	15751.2	1.0	**3.0986**	0.069	2.2	ug/L	201	Standard
	Pb	207	12955.9	1.0	**2.8124**	0.021	0.8	ug/L	177	Standard
	Pb	208	54935.0	0.7	**2.9710**	0.057	1.9	ug/L	722	Standard
	U	238	1866.8	0.6	**0.2674**	0.002	0.6	ug/L	8	Standard
>	Bi	209	447795.4	1.2				ug/L	444563	Standard

Sample ID: L1606145106 WG574521-02
Report Date/Time: Tuesday, July 05, 2016 12:03:09
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	55.0	45.5	**0.0381**	0.034	90.2	mg/L	38	Standard
K	39	46.7	6.2	**0.0812**	0.011	13.6	mg/L	33	Standard
Ca	43	58.3	44.0	**3.5989**	4.167	115.8	mg/L	48	Standard
Fe	54	1895.3	8.2	**1.4582**	0.136	9.3	mg/L	131	Standard
Fe	57	635.0	15.0	**1.5414**	0.262	17.0	mg/L	153	Standard
> **Sc-1**	45	25932.2	2.7				mg/L	26715	Standard
Cl	35	108583.1	2.1				ug/L	102686	Standard
Kr	83	1.3	43.3				ug/L	2	Standard
Br	81	1816.8	6.6				ug/L	1897	Standard
P	31	6998.3	6.1				ug/L	18491	Standard
S	34	3810.5	8.5				ug/L	3734	Standard
Sr	88	85.0	31.1				ug/L	47	Standard
C	12	120.0	14.4				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	50.0	69.3				mg/L	23	Standard
Dy	164	4605.9	7.9				mg/L	23	Standard
Ho-1	165	3028.6	4.1				mg/L	7	Standard
Er	166	2883.6	1.7				mg/L	13	Standard
I	127	29842.7	3.4				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		98.212	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.364	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145106 WG574521-02
Report Date/Time: Tuesday, July 05, 2016 12:03:09
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.763
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.727
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: L1606145106 WG574521-02
Report Date/Time: Tuesday, July 05, 2016 12:03:09
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, July 05, 2016 12:04:06
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	60838.6	1.2				ug/L	59517	Standard
	Be	9	44458.5	3.7	**50.0051**	2.323	4.6	ug/L	10	Standard
	Al	27	7478556.5	0.9	**48.0750**	0.433	0.9	ug/L	1872	Standard
	Sc	45	27386.4	1.8				ug/L	26715	Standard
	Ti	47	24816.6	1.5	**98.0648**	2.630	2.7	ug/L	54	Standard
	V	51	373142.7	0.9	**48.5325**	1.162	2.4	ug/L	1419	Standard
	Cr	52	381682.9	1.3	**49.5899**	1.284	2.6	ug/L	7059	Standard
	Cr	53	47979.2	0.5	**50.4138**	0.911	1.8	ug/L	398	Standard
	Mn	55	322593.6	1.6	**47.8905**	1.516	3.2	ug/L	824	Standard
	Co	59	347219.3	0.9	**48.1099**	1.187	2.5	ug/L	167	Standard
	Ni	60	99038.0	0.9	**49.6579**	0.386	0.8	ug/L	391	Standard
	Cu	65	98421.7	1.1	**49.3072**	1.146	2.3	ug/L	147	Standard
	Zn	66	57845.9	1.6	**49.2090**	1.532	3.1	ug/L	184	Standard
>	Ge	72	449181.6	1.5				ug/L	422998	Standard
	As	75	58659.0	1.3	**48.7545**	1.341	2.8	ug/L	-58	Standard
	Se	82	6000.6	1.4	**48.0056**	1.419	3.0	ug/L	15	Standard
	Se-1	77	3982.5	2.4	**48.8000**	1.663	3.4	ug/L	58	Standard
>	Ga	71	66.7	30.3				mg/L	113	Standard
	Rb	85	848.4	4.3				ug/L	27	Standard
	Y	89	388308.0	0.4				ug/L	383780	Standard
>	Rh	103	30.0	44.1				ug/L	18	Standard
	Mo	98	262276.3	0.9	**97.3335**	1.845	1.9	ug/L	31	Standard
	Ag	107	337726.1	1.6	**49.4532**	1.538	3.1	ug/L	84	Standard
	Cd	111	106537.8	1.8	**49.3683**	1.580	3.2	mg/L	4	Standard
	Cd	114	262775.5	1.0	**50.1687**	0.512	1.0	ug/L	15	Standard
>	In	115	522341.5	1.5				ug/L	515400	Standard
	Sn	118	289720.6	1.1	**49.2050**	1.308	2.7	ug/L	1043	Standard
	Sb	123	236187.5	1.8	**50.9109**	1.664	3.3	ug/L	227	Standard
	Ba	135	100057.6	0.8	**49.0989**	0.970	2.0	ug/L	31	Standard
	Ce	140	940.0	28.0				ug/L	120	Standard
>	Tb	159	807561.5	2.6				ug/L	785817	Standard
	Ho	165	11.7	89.2				ug/L	7	Standard
	Tl	203	403024.9	0.6	**49.1853**	1.084	2.2	ug/L	6	Standard
	Tl	205	356934.0	1.2	**50.0059**	1.362	2.7	ug/L	8	Standard
	Pb	206	253041.9	0.6	**49.6088**	1.160	2.3	ug/L	201	Standard
	Pb	207	229060.5	0.3	**49.6682**	0.925	1.9	ug/L	177	Standard
	Pb	208	939880.6	0.5	**50.7375**	1.077	2.1	ug/L	722	Standard
	U	238	345721.3	0.9	**49.1404**	1.276	2.6	ug/L	8	Standard
>	Bi	209	455564.3	1.7				ug/L	444563	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 12:06:23
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	3965.5	5.0	**4.8257**	0.272	5.6	mg/L	38	Standard
K	39	1005.0	6.7	**4.1666**	0.359	8.6	mg/L	33	Standard
Ca	43	63.3	9.1	**3.8479**	0.938	24.4	mg/L	48	Standard
Fe	54	6135.9	3.5	**4.7078**	0.127	2.7	mg/L	131	Standard
Fe	57	1786.8	2.7	**4.9214**	0.241	4.9	mg/L	153	Standard
Sc-1	45	27386.4	1.8				mg/L	26715	Standard
Cl	35	124070.0	1.6				ug/L	102686	Standard
Kr	83	2.7	21.7				ug/L	2	Standard
Br	81	1790.1	12.7				ug/L	1897	Standard
P	31	20143.4	2.1				ug/L	18491	Standard
S	34	4897.5	4.4				ug/L	3734	Standard
Sr	88	80.0	21.7				ug/L	47	Standard
C	12	53.3	60.3				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	16.7	34.6				mg/L	23	Standard
Dy	164	15.7	36.8				mg/L	23	Standard
Ho-1	165	11.7	89.2				mg/L	7	Standard
Er	166	20.0	0.0				mg/L	13	Standard
I	127	15040.8	3.6				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	100.010		
Al	27	96.150		
Sc	45			
Ti	47	98.065		
V	51	97.065		
Cr	52	99.180		
Cr	53			
Mn	55	95.781		
Co	59	96.220		
Ni	60	99.316		
Cu	65	98.614		
Zn	66	98.418		
Ge	72		106.190	
As	75	97.509		
Se	82	96.011		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 12:06:23
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
Rh	103			
Mo	98	97.333		
Ag	107	98.906		
Cd	111	98.737		
Cd	114			
In	115			101.347
Sn	118	98.410		
Sb	123	101.822		
Ba	135	98.198		
Ce	140			
Tb	159			
Ho	165			
Tl	203	98.371		
Tl	205			
Pb	206			
Pb	207			
Pb	208	101.475		
U	238	98.281		
Bi	209			102.475
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 12:06:23
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, July 05, 2016 12:07:17
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	61857.7	0.8				ug/L	59517	Standard
	Be	9	23.3	32.7	**-0.0053**	0.008	157.3	ug/L	10	Standard
	Al	27	2840.3	34.2	**0.0055**	0.006	108.9	ug/L	1872	Standard
	Sc	45	27840.6	2.2				ug/L	26715	Standard
	Ti	47	28.3	35.5	**-0.1363**	0.042	30.8	ug/L	54	Standard
	V	51	1397.2	2.3	**-0.0010**	0.005	459.1	ug/L	1419	Standard
	Cr	52	6780.5	1.1	**-0.0969**	0.030	31.3	ug/L	7059	Standard
	Cr	53	1530.1	2.9	**1.1752**	0.029	2.5	ug/L	398	Standard
	Mn	55	835.7	22.7	**-0.0088**	0.031	346.9	ug/L	824	Standard
	Co	59	241.0	5.7	**0.0076**	0.003	35.3	ug/L	167	Standard
	Ni	60	285.7	9.1	**-0.0592**	0.012	21.0	ug/L	391	Standard
	Cu	65	133.3	2.2	**-0.0136**	0.002	16.8	ug/L	147	Standard
	Zn	66	252.0	10.0	**0.0514**	0.026	50.7	ug/L	184	Standard
>	Ge	72	459148.3	2.5				ug/L	422998	Standard
	As	75	-67.2	62.8	**-0.0030**	0.034	1129.6	ug/L	-58	Standard
	Se	82	15.4	38.4	**-0.0850**	0.045	52.5	ug/L	15	Standard
	Se-1	77	99.7	5.5	**0.5426**	0.061	11.3	ug/L	58	Standard
>	Ga	71	61.7	33.8				mg/L	113	Standard
	Rb	85	38.3	32.8				ug/L	27	Standard
	Y	89	400704.7	0.8				ug/L	383780	Standard
>	Rh	103	5.0	0.0				ug/L	18	Standard
	Mo	98	193.4	7.3	**0.0598**	0.006	10.8	ug/L	31	Standard
	Ag	107	117.3	17.2	**0.0045**	0.003	70.6	ug/L	84	Standard
	Cd	111	16.0	42.5	**0.0030**	0.003	106.1	mg/L	4	Standard
	Cd	114	42.8	62.4	**0.0033**	0.005	152.1	ug/L	15	Standard
>	In	115	546361.8	2.1				ug/L	515400	Standard
	Sn	118	1136.7	26.0	**0.0036**	0.044	1221.9	ug/L	1043	Standard
	Sb	123	943.1	28.8	**0.1773**	0.052	29.4	ug/L	227	Standard
	Ba	135	71.7	32.3	**0.0101**	0.012	115.1	ug/L	31	Standard
	Ce	140	466.7	35.0				ug/L	120	Standard
>	Tb	159	826410.9	2.5				ug/L	785817	Standard
	Ho	165	8.3	34.6				ug/L	7	Standard
	Tl	203	76.3	25.3	**0.0054**	0.003	46.7	ug/L	6	Standard
	Tl	205	55.0	15.7	**0.0006**	0.001	215.3	ug/L	8	Standard
	Pb	206	285.0	11.5	**0.0082**	0.008	92.1	ug/L	201	Standard
	Pb	207	244.3	6.4	**0.0031**	0.005	146.8	ug/L	177	Standard
	Pb	208	1017.7	17.3	**0.0044**	0.011	242.9	ug/L	722	Standard
	U	238	60.3	24.4	**0.0058**	0.002	38.8	ug/L	8	Standard
>	Bi	209	474436.7	2.5				ug/L	444563	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 12:09:34
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	23.3	44.6	**-0.0056**	0.012	214.0	mg/L	38	Standard
K	39	30.0	76.4	**-0.0044**	0.094	2156.6	mg/L	33	Standard
Ca	43	30.0	28.9	**-1.3278**	1.429	107.6	mg/L	48	Standard
Fe	54	129.2	16.9	**-0.0167**	0.016	96.0	mg/L	131	Standard
Fe	57	226.7	12.5	**0.1931**	0.097	50.1	mg/L	153	Standard
Sc-1	45	27840.6	2.2				mg/L	26715	Standard
Cl	35	120675.4	1.5				ug/L	102686	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	1633.4	7.0				ug/L	1897	Standard
P	31	20031.6	3.6				ug/L	18491	Standard
S	34	4719.1	2.8				ug/L	3734	Standard
Sr	88	66.7	37.0				ug/L	47	Standard
C	12	66.7	31.2				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	16.7	91.7				mg/L	23	Standard
Dy	164	22.9	91.6				mg/L	23	Standard
Ho-1	165	8.3	34.6				mg/L	7	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	9609.7	3.4				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		108.546	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 12:09:34
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		106.007
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		106.720
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 12:09:34
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145107
Sample Date/Time: Tuesday, July 05, 2016 12:10:30
Number of Replicates: 3
Autosampler Position: 235
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	59463.2	2.0				ug/L	59517	Standard
	Be	9	271.7	10.1	**0.2818**	0.034	12.0	ug/L	10	Standard
	Al	27	563585.4	1.6	**3.6964**	0.112	3.0	ug/L	1872	Standard
	Sc	45	25695.1	4.1				ug/L	26715	Standard
	Ti	47	9216.8	0.7	**38.2822**	0.331	0.9	ug/L	54	Standard
	V	51	30052.2	0.8	**3.9605**	0.059	1.5	ug/L	1419	Standard
	Cr	52	32999.6	0.8	**3.6372**	0.036	1.0	ug/L	7059	Standard
	Cr	53	5007.5	5.3	**5.1872**	0.291	5.6	ug/L	398	Standard
	Mn	55	126812.5	1.1	**19.7882**	0.336	1.7	ug/L	824	Standard
	Co	59	4650.7	2.1	**0.6552**	0.013	2.0	ug/L	167	Standard
	Ni	60	4558.7	2.3	**2.2223**	0.061	2.8	ug/L	391	Standard
	Cu	65	5570.0	2.4	**2.8700**	0.046	1.6	ug/L	147	Standard
	Zn	66	19959.8	1.3	**17.8147**	0.249	1.4	ug/L	184	Standard
>	**Ge**	72	425585.2	0.8				ug/L	422998	Standard
	As	75	6588.1	2.5	**5.8230**	0.127	2.2	ug/L	-58	Standard
	Se	82	41.3	9.5	**0.1439**	0.033	22.7	ug/L	15	Standard
	Se-1	77	123.7	16.8	**0.9511**	0.259	27.2	ug/L	58	Standard
>	**Ga**	71	2348.5	2.9				mg/L	113	Standard
	Rb	85	18669.8	1.9				ug/L	27	Standard
	Y	89	385954.3	1.2				ug/L	383780	Standard
>	**Rh**	103	13.3	57.3				ug/L	18	Standard
	Mo	98	1308.2	3.9	**0.5002**	0.028	5.6	ug/L	31	Standard
	Ag	107	158.0	10.6	**0.0123**	0.003	23.1	ug/L	84	Standard
	Cd	111	419.5	1.3	**0.1996**	0.006	3.1	mg/L	4	Standard
	Cd	114	991.1	7.1	**0.1939**	0.017	8.5	ug/L	15	Standard
>	**In**	115	498311.6	1.9				ug/L	515400	Standard
	Sn	118	643.3	30.4	**-0.0660**	0.035	53.4	ug/L	1043	Standard
	Sb	123	717.3	29.3	**0.1456**	0.047	32.4	ug/L	227	Standard
	Ba	135	21850.8	2.4	**11.2195**	0.221	2.0	ug/L	31	Standard
	Ce	140	70835.2	0.9				ug/L	120	Standard
>	**Tb**	159	773727.5	1.8				ug/L	785817	Standard
	Ho	165	1485.1	8.6				ug/L	7	Standard
	Tl	203	2063.5	3.9	**0.2523**	0.006	2.2	ug/L	6	Standard
	Tl	205	1733.4	7.2	**0.2399**	0.013	5.6	ug/L	8	Standard
	Pb	206	29512.4	0.5	**5.8409**	0.134	2.3	ug/L	201	Standard
	Pb	207	24603.3	0.8	**5.3801**	0.114	2.1	ug/L	177	Standard
	Pb	208	103716.4	1.5	**5.6481**	0.155	2.8	ug/L	722	Standard
	U	238	1387.7	4.4	**0.1979**	0.007	3.5	ug/L	8	Standard
>	**Bi**	209	448200.2	2.0				ug/L	444563	Standard

Sample ID: L1606145107
Report Date/Time: Tuesday, July 05, 2016 12:12:47
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	50.0	10.0	**0.0319**	0.009	27.3	mg/L	38	Standard
K	39	66.7	8.7	**0.1752**	0.038	21.6	mg/L	33	Standard
Ca	43	35.0	51.5	**-0.1982**	2.805	1415.4	mg/L	48	Standard
Fe	54	4006.8	3.7	**3.2418**	0.015	0.5	mg/L	131	Standard
Fe	57	1298.4	5.5	**3.7102**	0.376	10.1	mg/L	153	Standard
> **Sc-1**	45	25695.1	4.1				mg/L	26715	Standard
Cl	35	106066.0	0.7				ug/L	102686	Standard
Kr	83	2.0	132.3				ug/L	2	Standard
Br	81	1646.8	11.5				ug/L	1897	Standard
P	31	7333.4	1.7				ug/L	18491	Standard
S	34	4165.6	0.6				ug/L	3734	Standard
Sr	88	95.0	29.3				ug/L	47	Standard
C	12	50.0	105.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	6.7	86.6				mg/L	23	Standard
Dy	164	2350.0	8.7				mg/L	23	Standard
Ho-1	165	1485.1	8.6				mg/L	7	Standard
Er	166	1396.7	6.1				mg/L	13	Standard
I	127	27097.6	2.5				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		99.910	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		100.612	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145107
Report Date/Time: Tuesday, July 05, 2016 12:12:47
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		96.684
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		100.818
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145108
Sample Date/Time: Tuesday, July 05, 2016 12:13:42
Number of Replicates: 3
Autosampler Position: 236
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	56130.6	4.3				ug/L	59517	Standard
	Be	9	390.0	2.6	**0.4454**	0.033	7.3	ug/L	10	Standard
	Al	27	714616.0	1.9	**4.9765**	0.305	6.1	ug/L	1872	Standard
	Sc	45	24990.6	1.8				ug/L	26715	Standard
	Ti	47	10313.5	2.7	**44.3310**	0.816	1.8	ug/L	54	Standard
	V	51	31216.7	3.2	**4.2666**	0.094	2.2	ug/L	1419	Standard
	Cr	52	51017.4	1.8	**6.3994**	0.132	2.1	ug/L	7059	Standard
	Cr	53	7175.0	5.9	**7.8784**	0.311	4.0	ug/L	398	Standard
	Mn	55	144939.4	1.4	**23.4138**	0.562	2.4	ug/L	824	Standard
	Co	59	5125.2	1.4	**0.7502**	0.006	0.8	ug/L	167	Standard
	Ni	60	5278.3	2.9	**2.6996**	0.049	1.8	ug/L	391	Standard
	Cu	65	5093.2	3.7	**2.7090**	0.062	2.3	ug/L	147	Standard
	Zn	66	20594.0	1.5	**19.0178**	0.238	1.3	ug/L	184	Standard
>	Ge	72	411597.6	2.1				ug/L	422998	Standard
	As	75	6186.0	2.4	**5.6554**	0.089	1.6	ug/L	-58	Standard
	Se	82	33.8	9.8	**0.0904**	0.033	36.9	ug/L	15	Standard
	Se-1	77	117.7	7.9	**0.9285**	0.158	17.0	ug/L	58	Standard
>	Ga	71	2496.9	1.1				mg/L	113	Standard
	Rb	85	21094.7	1.3				ug/L	27	Standard
	Y	89	375854.5	2.2				ug/L	383780	Standard
>	Rh	103	15.0	57.7				ug/L	18	Standard
	Mo	98	1308.0	3.2	**0.5065**	0.011	2.2	ug/L	31	Standard
	Ag	107	178.3	7.1	**0.0157**	0.002	11.1	ug/L	84	Standard
	Cd	111	460.5	9.7	**0.2223**	0.019	8.6	mg/L	4	Standard
	Cd	114	1153.3	11.3	**0.2291**	0.024	10.6	ug/L	15	Standard
>	In	115	491828.9	1.4				ug/L	515400	Standard
	Sn	118	501.7	18.7	**-0.0903**	0.016	17.3	ug/L	1043	Standard
	Sb	123	553.0	26.0	**0.1100**	0.031	28.5	ug/L	227	Standard
	Ba	135	25357.2	2.2	**13.1943**	0.148	1.1	ug/L	31	Standard
	Ce	140	83605.1	1.4				ug/L	120	Standard
>	Tb	159	755464.3	0.3				ug/L	785817	Standard
	Ho	165	1738.4	1.8				ug/L	7	Standard
	Tl	203	2133.2	1.5	**0.2671**	0.002	0.9	ug/L	6	Standard
	Tl	205	1770.1	6.4	**0.2511**	0.016	6.3	ug/L	8	Standard
	Pb	206	27937.7	1.6	**5.6556**	0.062	1.1	ug/L	201	Standard
	Pb	207	23353.0	1.5	**5.2233**	0.017	0.3	ug/L	177	Standard
	Pb	208	97617.9	2.7	**5.4374**	0.163	3.0	ug/L	722	Standard
	U	238	1483.4	1.0	**0.2168**	0.005	2.4	ug/L	8	Standard
>	Bi	209	437967.3	1.5				ug/L	444563	Standard

Sample ID: L1606145108
Report Date/Time: Tuesday, July 05, 2016 12:15:59
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	60.0	14.4	**0.0469**	0.012	24.6	mg/L	38	Standard
K	39	86.7	13.3	**0.2770**	0.061	22.1	mg/L	33	Standard
Ca	43	38.3	19.9	**0.5772**	1.345	233.0	mg/L	48	Standard
Fe	54	4538.8	2.5	**3.7943**	0.055	1.4	mg/L	131	Standard
Fe	57	1466.7	10.1	**4.3719**	0.419	9.6	mg/L	153	Standard
> **Sc-1**	45	24990.6	1.8				mg/L	26715	Standard
Cl	35	105965.3	1.5				ug/L	102686	Standard
Kr	83	2.3	49.5				ug/L	2	Standard
Br	81	1653.4	9.8				ug/L	1897	Standard
P	31	7376.8	0.4				ug/L	18491	Standard
S	34	4207.3	3.6				ug/L	3734	Standard
Sr	88	85.0	29.4				ug/L	47	Standard
C	12	80.0	12.5				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	50.0				mg/L	23	Standard
Dy	164	2392.9	8.1				mg/L	23	Standard
Ho-1	165	1738.4	1.8				mg/L	7	Standard
Er	166	1543.4	17.7				mg/L	13	Standard
I	127	28022.6	5.3				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.311	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		97.305	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145108
Report Date/Time: Tuesday, July 05, 2016 12:15:59
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.427
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		98.516
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145109S WG574521-07
Sample Date/Time: Tuesday, July 05, 2016 12:16:54
Number of Replicates: 3
Autosampler Position: 237
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	56190.8	4.9				ug/L	59517	Standard
	Be	9	2413.5	4.5	**2.9148**	0.218	7.5	ug/L	10	Standard
	Al	27	371125.2	2.4	**2.5768**	0.177	6.9	ug/L	1872	Standard
	Sc	45	26199.3	1.2				ug/L	26715	Standard
	Ti	47	2656.2	3.1	**10.9848**	0.609	5.5	ug/L	54	Standard
	V	51	72353.8	1.8	**9.9011**	0.444	4.5	ug/L	1419	Standard
	Cr	52	39238.4	2.3	**4.5718**	0.251	5.5	ug/L	7059	Standard
	Cr	53	5817.8	5.6	**6.1713**	0.535	8.7	ug/L	398	Standard
	Mn	55	194686.5	0.9	**30.7906**	1.037	3.4	ug/L	824	Standard
	Co	59	20292.2	0.6	**2.9764**	0.090	3.0	ug/L	167	Standard
	Ni	60	10310.8	3.5	**5.3414**	0.313	5.9	ug/L	391	Standard
	Cu	65	16805.3	3.2	**8.9219**	0.495	5.5	ug/L	147	Standard
	Zn	66	16269.4	2.6	**14.6606**	0.774	5.3	ug/L	184	Standard
>	Ge	72	421107.8	2.7				ug/L	422998	Standard
	As	75	4858.7	2.3	**4.3571**	0.207	4.7	ug/L	-58	Standard
	Se	82	413.1	4.8	**3.3379**	0.258	7.7	ug/L	15	Standard
	Se-1	77	370.0	5.4	**4.2398**	0.392	9.3	ug/L	58	Standard
>	Ga	71	1511.7	8.6				mg/L	113	Standard
	Rb	85	5796.1	4.4				ug/L	27	Standard
	Y	89	409247.1	1.2				ug/L	383780	Standard
>	Rh	103	13.3	21.7				ug/L	18	Standard
	Mo	98	1013.4	3.4	**0.3905**	0.021	5.5	ug/L	31	Standard
	Ag	107	16993.5	2.1	**2.6302**	0.114	4.3	ug/L	84	Standard
	Cd	111	6180.4	3.0	**3.0365**	0.151	5.0	mg/L	4	Standard
	Cd	114	14740.9	0.9	**2.9837**	0.103	3.4	ug/L	15	Standard
>	In	115	492243.2	2.7				ug/L	515400	Standard
	Sn	118	545.0	27.7	**-0.0819**	0.030	36.9	ug/L	1043	Standard
	Sb	123	1618.1	8.4	**0.3542**	0.036	10.2	ug/L	227	Standard
	Ba	135	34944.9	1.6	**18.1908**	0.739	4.1	ug/L	31	Standard
	Ce	140	160735.2	2.9				ug/L	120	Standard
>	Tb	159	765593.4	3.7				ug/L	785817	Standard
	Ho	165	3587.1	5.4				ug/L	7	Standard
	Tl	203	20903.4	1.2	**2.6653**	0.086	3.2	ug/L	6	Standard
	Tl	205	17982.3	0.6	**2.6285**	0.058	2.2	ug/L	8	Standard
	Pb	206	32329.5	2.4	**6.5920**	0.298	4.5	ug/L	201	Standard
	Pb	207	27598.8	3.1	**6.2208**	0.340	5.5	ug/L	177	Standard
	Pb	208	115863.4	1.6	**6.5015**	0.254	3.9	ug/L	722	Standard
	U	238	19080.3	2.2	**2.8353**	0.131	4.6	ug/L	8	Standard
>	Bi	209	435586.5	2.4				ug/L	444563	Standard

Sample ID: L1606145109S WG574521-07
Report Date/Time: Tuesday, July 05, 2016 12:19:11
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**19.6496**	34.025	173.2	mg/L	0	Standard
Mg	24	65.0	35.3	**0.0496**	0.029	59.3	mg/L	38	Standard
K	39	51.7	31.1	**0.1009**	0.069	68.2	mg/L	33	Standard
Ca	43	33.3	37.7	**-0.5368**	2.007	373.9	mg/L	48	Standard
Fe	54	2186.8	2.4	**1.6814**	0.065	3.9	mg/L	131	Standard
Fe	57	733.4	12.3	**1.8375**	0.311	16.9	mg/L	153	Standard
> **Sc-1**	45	26199.3	1.2				mg/L	26715	Standard
Cl	35	108414.5	1.1				ug/L	102686	Standard
Kr	83	2.0	50.0				ug/L	2	Standard
Br	81	2523.5	7.7				ug/L	1897	Standard
P	31	7423.5	1.0				ug/L	18491	Standard
S	34	4045.5	5.1				ug/L	3734	Standard
Sr	88	93.3	15.5				ug/L	47	Standard
C	12	136.7	8.4				mg/L	63	Standard
N	14	3.3	173.2				mg/L	0	Standard
Hg	202	43.3	35.3				mg/L	23	Standard
Dy	164	5441.4	1.3				mg/L	23	Standard
Ho-1	165	3587.1	5.4				mg/L	7	Standard
Er	166	3540.4	2.6				mg/L	13	Standard
I	127	33304.9	1.8				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		94.412	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		99.553	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145109S WG574521-07
Report Date/Time: Tuesday, July 05, 2016 12:19:11
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		95.507
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		97.981
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: L1606145109S WG574521-07
Report Date/Time: Tuesday, July 05, 2016 12:19:11
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606145110SD WG574521-08
Sample Date/Time: Tuesday, July 05, 2016 12:20:05
Number of Replicates: 3
Autosampler Position: 238
Sample Description: 10
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	Li	6	54180.1	3.9				ug/L	59517	Standard
	Be	9	2170.2	6.2	**2.7136**	0.200	7.4	ug/L	10	Standard
	Al	27	298363.5	2.5	**2.1427**	0.042	2.0	ug/L	1872	Standard
	Sc	45	25074.1	0.3				ug/L	26715	Standard
	Ti	47	1478.7	2.7	**6.1802**	0.101	1.6	ug/L	54	Standard
	V	51	60367.5	1.2	**8.4684**	0.144	1.7	ug/L	1419	Standard
	Cr	52	33902.9	0.8	**3.9533**	0.111	2.8	ug/L	7059	Standard
	Cr	53	5007.5	1.8	**5.4111**	0.136	2.5	ug/L	398	Standard
	Mn	55	162958.4	1.0	**26.4853**	0.261	1.0	ug/L	824	Standard
	Co	59	18000.0	1.4	**2.7132**	0.050	1.8	ug/L	167	Standard
	Ni	60	8905.3	1.3	**4.7202**	0.058	1.2	ug/L	391	Standard
	Cu	65	14460.9	1.0	**7.8846**	0.181	2.3	ug/L	147	Standard
	Zn	66	12677.6	1.2	**11.7127**	0.193	1.6	ug/L	184	Standard
>	Ge	72	409283.0	1.5				ug/L	422998	Standard
	As	75	3802.4	1.1	**3.5163**	0.084	2.4	ug/L	-58	Standard
	Se	82	363.9	1.2	**3.0022**	0.084	2.8	ug/L	15	Standard
	Se-1	77	342.0	2.1	**3.9927**	0.069	1.7	ug/L	58	Standard
>	Ga	71	1035.0	4.6				mg/L	113	Standard
	Rb	85	2341.8	7.3				ug/L	27	Standard
	Y	89	387313.8	1.7				ug/L	383780	Standard
>	Rh	103	11.7	24.7				ug/L	18	Standard
	Mo	98	998.1	4.5	**0.3934**	0.015	3.9	ug/L	31	Standard
	Ag	107	14925.7	0.6	**2.3624**	0.029	1.2	ug/L	84	Standard
	Cd	111	5660.9	2.7	**2.8461**	0.121	4.3	mg/L	4	Standard
	Cd	114	13763.6	2.5	**2.8510**	0.106	3.7	ug/L	15	Standard
>	In	115	480835.3	1.7				ug/L	515400	Standard
	Sn	118	533.3	22.2	**-0.0824**	0.020	24.5	ug/L	1043	Standard
	Sb	123	3307.7	1.1	**0.7582**	0.005	0.6	ug/L	227	Standard
	Ba	135	29372.8	0.5	**15.6422**	0.331	2.1	ug/L	31	Standard
	Ce	140	135459.5	0.9				ug/L	120	Standard
>	Tb	159	748567.3	2.0				ug/L	785817	Standard
	Ho	165	3172.0	2.9				ug/L	7	Standard
	Tl	203	18739.2	0.5	**2.4241**	0.040	1.7	ug/L	6	Standard
	Tl	205	16330.5	1.2	**2.4221**	0.055	2.3	ug/L	8	Standard
	Pb	206	29114.6	1.1	**6.0183**	0.074	1.2	ug/L	201	Standard
	Pb	207	24724.2	0.6	**5.6484**	0.087	1.5	ug/L	177	Standard
	Pb	208	103982.3	1.0	**5.9154**	0.098	1.6	ug/L	722	Standard
	U	238	17333.6	0.9	**2.6126**	0.027	1.0	ug/L	8	Standard
>	Bi	209	429156.7	1.9				ug/L	444563	Standard

Sample ID: L1606145110SD WG574521-08
Report Date/Time: Tuesday, July 05, 2016 12:22:22
Page 1

Approved: July 06, 2016

Na	23	1.7	173.2	**20.8696**	36.139	173.2	mg/L	0	Standard
Mg	24	55.0	68.6	**0.0400**	0.051	126.7	mg/L	38	Standard
K	39	21.7	58.1	**-0.0281**	0.059	208.8	mg/L	33	Standard
Ca	43	45.0	33.3	**1.6609**	2.492	150.0	mg/L	48	Standard
Fe	54	1724.9	4.1	**1.3647**	0.058	4.3	mg/L	131	Standard
Fe	57	648.3	11.4	**1.6582**	0.244	14.7	mg/L	153	Standard
> **Sc-1**	45	25074.1	0.3				mg/L	26715	Standard
Cl	35	110353.2	2.0				ug/L	102686	Standard
Kr	83	2.3	24.7				ug/L	2	Standard
Br	81	2023.5	8.7				ug/L	1897	Standard
P	31	7188.4	0.9				ug/L	18491	Standard
S	34	4012.2	5.7				ug/L	3734	Standard
Sr	88	96.7	10.8				ug/L	47	Standard
C	12	146.7	19.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	53.3	39.0				mg/L	23	Standard
Dy	164	4874.9	7.0				mg/L	23	Standard
Ho-1	165	3172.0	2.9				mg/L	7	Standard
Er	166	3047.0	8.4				mg/L	13	Standard
I	127	30377.1	0.5				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		91.033	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		96.758	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606145110SD WG574521-08
Report Date/Time: Tuesday, July 05, 2016 12:22:22
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		93.294
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		96.535
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: L1606145110SD WG574521-08
Report Date/Time: Tuesday, July 05, 2016 12:22:22
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 6
Sample Date/Time: Tuesday, July 05, 2016 12:23:18
Number of Replicates: 3
Autosampler Position: 101
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	61936.3	1.0				ug/L	59517	Standard
	Be	9	44060.7	4.1	**48.6681**	2.073	4.3	ug/L	10	Standard
	Al	27	7421017.2	0.9	**46.8624**	0.837	1.8	ug/L	1872	Standard
	Sc	45	27840.6	0.8				ug/L	26715	Standard
	Ti	47	24780.2	1.3	**94.1129**	1.402	1.5	ug/L	54	Standard
	V	51	374015.7	2.5	**46.7546**	1.361	2.9	ug/L	1419	Standard
	Cr	52	381449.2	1.0	**47.5991**	0.617	1.3	ug/L	7059	Standard
	Cr	53	48109.6	1.3	**48.5792**	0.817	1.7	ug/L	398	Standard
	Mn	55	323449.1	1.4	**46.1470**	0.763	1.7	ug/L	824	Standard
	Co	59	353958.9	1.7	**47.1411**	0.942	2.0	ug/L	167	Standard
	Ni	60	99916.4	1.4	**48.1579**	0.828	1.7	ug/L	391	Standard
	Cu	65	98273.1	2.2	**47.3222**	1.174	2.5	ug/L	147	Standard
	Zn	66	58263.5	2.1	**47.6343**	1.129	2.4	ug/L	184	Standard
>	**Ge**	72	467206.7	0.4				ug/L	422998	Standard
	As	75	58897.6	1.9	**47.0555**	1.055	2.2	ug/L	-58	Standard
	Se	82	6005.7	2.7	**46.1739**	1.361	2.9	ug/L	15	Standard
	Se-1	77	4137.9	0.2	**48.7349**	0.165	0.3	ug/L	58	Standard
>	**Ga**	71	86.7	12.0				mg/L	113	Standard
	Rb	85	898.4	11.7				ug/L	27	Standard
	Y	89	406260.6	1.7				ug/L	383780	Standard
>	**Rh**	103	20.0	25.0				ug/L	18	Standard
	Mo	98	265996.1	0.8	**93.3198**	1.560	1.7	ug/L	31	Standard
	Ag	107	339646.5	1.4	**47.0210**	1.585	3.4	ug/L	84	Standard
	Cd	111	108529.1	1.3	**47.5497**	1.671	3.5	mg/L	4	Standard
	Cd	114	269565.9	1.7	**48.6723**	1.830	3.8	ug/L	15	Standard
>	**In**	115	552561.7	2.2				ug/L	515400	Standard
	Sn	118	291910.6	1.5	**46.8625**	1.419	3.0	ug/L	1043	Standard
	Sb	123	238681.8	1.8	**48.6440**	1.842	3.8	ug/L	227	Standard
	Ba	135	100767.3	0.6	**46.7520**	1.319	2.8	ug/L	31	Standard
	Ce	140	696.7	32.8				ug/L	120	Standard
>	**Tb**	159	836664.0	1.8				ug/L	785817	Standard
	Ho	165	10.0	50.0				ug/L	7	Standard
	Tl	203	408219.9	0.5	**47.9579**	0.937	2.0	ug/L	6	Standard
	Tl	205	353779.4	0.6	**47.7113**	1.066	2.2	ug/L	8	Standard
	Pb	206	254575.9	0.6	**48.0435**	1.017	2.1	ug/L	201	Standard
	Pb	207	231354.7	0.8	**48.2933**	1.083	2.2	ug/L	177	Standard
	Pb	208	942309.5	0.6	**48.9650**	0.796	1.6	ug/L	722	Standard
	U	238	347285.7	0.5	**47.5170**	0.991	2.1	ug/L	8	Standard
>	**Bi**	209	473225.0	1.6				ug/L	444563	Standard

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 12:25:35
Page 1

Approved: July 06, 2016

Na	23	3.3	173.2	**37.3364**	64.660	173.2	mg/L	0	Standard
Mg	24	3865.5	1.1	**4.6244**	0.043	0.9	mg/L	38	Standard
K	39	880.0	4.1	**3.5668**	0.125	3.5	mg/L	33	Standard
Ca	43	60.0	38.2	**3.1652**	3.408	107.7	mg/L	48	Standard
Fe	54	6324.7	4.0	**4.7753**	0.185	3.9	mg/L	131	Standard
Fe	57	1893.5	4.7	**5.1469**	0.235	4.6	mg/L	153	Standard
Sc-1	45	27840.6	0.8				mg/L	26715	Standard
Cl	35	126370.8	0.7				ug/L	102686	Standard
Kr	83	1.3	114.6				ug/L	2	Standard
Br	81	1893.5	8.0				ug/L	1897	Standard
P	31	21593.7	1.8				ug/L	18491	Standard
S	34	4702.4	6.9				ug/L	3734	Standard
Sr	88	61.7	20.4				ug/L	47	Standard
C	12	43.3	66.6				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	86.6				mg/L	23	Standard
Dy	164	19.4	1.4				mg/L	23	Standard
Ho-1	165	10.0	50.0				mg/L	7	Standard
Er	166	13.3	43.3				mg/L	13	Standard
I	127	14653.8	2.4				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9	97.336		
Al	27	93.725		
Sc	45			
Ti	47	94.113		
V	51	93.509		
Cr	52	95.198		
Cr	53			
Mn	55	92.294		
Co	59	94.282		
Ni	60	96.316		
Cu	65	94.644		
Zn	66	95.269		
Ge	72		110.451	
As	75	94.111		
Se	82	92.348		
Se-1	77			
Ga	71			

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 12:25:35
Page 2

Approved: July 06, 2016

Rb	85			
Y	89			
> Rh	103			
Mo	98	93.320		
Ag	107	94.042		
Cd	111	95.099		
Cd	114			
> In	115			107.210
Sn	118	93.725		
Sb	123	97.288		
Ba	135	93.504		
Ce	140			
> Tb	159			
Ho	165			
Tl	203	95.916		
Tl	205			
Pb	206			
Pb	207			
Pb	208	97.930		
U	238	95.034		
> Bi	209			106.447
Na	23			
Mg	24			
K	39			
Ca	43			
Fe	54			
Fe	57			
> Sc-1	45			
Cl	35			
Kr	83			
Br	81			
P	31			
S	34			
Sr	88			
C	12			
N	14			
Hg	202			
Dy	164			
Ho-1	165			
Er	166			
I	127			

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 6
Report Date/Time: Tuesday, July 05, 2016 12:25:35
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: QC Std 7
Sample Date/Time: Tuesday, July 05, 2016 12:26:29
Number of Replicates: 3
Autosampler Position: 102
Sample Description:
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	63303.6	1.3				ug/L	59517	Standard
	Be	9	18.3	31.5	**-0.0113**	0.006	52.7	ug/L	10	Standard
	Al	27	2235.2	14.5	**0.0014**	0.002	134.0	ug/L	1872	Standard
	Sc	45	28122.8	2.8				ug/L	26715	Standard
	Ti	47	35.3	21.2	**-0.1115**	0.029	25.7	ug/L	54	Standard
	V	51	1439.5	6.3	**0.0020**	0.010	505.5	ug/L	1419	Standard
	Cr	52	6652.8	0.6	**-0.1246**	0.005	3.6	ug/L	7059	Standard
	Cr	53	1205.0	5.1	**0.8226**	0.054	6.5	ug/L	398	Standard
	Mn	55	652.7	4.2	**-0.0371**	0.004	11.7	ug/L	824	Standard
	Co	59	238.3	13.0	**0.0068**	0.004	58.9	ug/L	167	Standard
	Ni	60	288.0	2.5	**-0.0598**	0.004	7.2	ug/L	391	Standard
	Cu	65	129.3	6.7	**-0.0163**	0.004	26.3	ug/L	147	Standard
	Zn	66	468.0	81.2	**0.2258**	0.312	138.0	ug/L	184	Standard
>	**Ge**	72	464917.0	0.8				ug/L	422998	Standard
	As	75	-45.4	55.4	**0.0153**	0.020	130.4	ug/L	-58	Standard
	Se	82	16.5	14.7	**-0.0781**	0.020	25.1	ug/L	15	Standard
	Se-1	77	95.3	12.3	**0.4748**	0.132	27.9	ug/L	58	Standard
>	**Ga**	71	45.0	11.1				mg/L	113	Standard
	Rb	85	6.7	114.6				ug/L	27	Standard
	Y	89	398203.8	1.4				ug/L	383780	Standard
>	**Rh**	103	16.7	75.5				ug/L	18	Standard
	Mo	98	228.0	6.8	**0.0700**	0.006	7.9	ug/L	31	Standard
	Ag	107	151.7	7.6	**0.0087**	0.001	16.7	ug/L	84	Standard
	Cd	111	23.3	21.2	**0.0059**	0.002	35.0	mg/L	4	Standard
	Cd	114	47.2	54.5	**0.0038**	0.005	118.3	ug/L	15	Standard
>	**In**	115	560315.0	0.7				ug/L	515400	Standard
	Sn	118	1160.0	23.8	**0.0035**	0.045	1275.6	ug/L	1043	Standard
	Sb	123	862.8	32.5	**0.1573**	0.057	36.4	ug/L	227	Standard
	Ba	135	48.3	39.2	**-0.0016**	0.009	518.6	ug/L	31	Standard
	Ce	140	270.0	27.3				ug/L	120	Standard
>	**Tb**	159	838378.6	1.5				ug/L	785817	Standard
	Ho	165	8.3	69.3				ug/L	7	Standard
	Tl	203	88.7	16.3	**0.0066**	0.002	26.0	ug/L	6	Standard
	Tl	205	103.3	19.6	**0.0069**	0.003	39.7	ug/L	8	Standard
	Pb	206	274.3	2.4	**0.0053**	0.002	28.7	ug/L	201	Standard
	Pb	207	254.3	11.1	**0.0042**	0.006	142.5	ug/L	177	Standard
	Pb	208	1025.0	7.9	**0.0037**	0.004	117.5	ug/L	722	Standard
	U	238	84.0	17.2	**0.0088**	0.002	22.6	ug/L	8	Standard
>	**Bi**	209	482466.6	0.6				ug/L	444563	Standard

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 12:28:46
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	31.7	39.7	**0.0042**	0.015	355.9	mg/L	38	Standard
K	39	13.3	21.7	**-0.0736**	0.013	17.7	mg/L	33	Standard
Ca	43	35.0	37.8	**-0.6810**	1.859	273.0	mg/L	48	Standard
Fe	54	177.9	22.8	**0.0196**	0.030	154.7	mg/L	131	Standard
Fe	57	193.3	5.4	**0.0870**	0.025	28.9	mg/L	153	Standard
Sc-1	45	28122.8	2.8				mg/L	26715	Standard
Cl	35	120247.5	2.4				ug/L	102686	Standard
Kr	83	1.0	100.0				ug/L	2	Standard
Br	81	1830.1	19.2				ug/L	1897	Standard
P	31	19964.8	2.8				ug/L	18491	Standard
S	34	4852.4	0.8				ug/L	3734	Standard
Sr	88	66.7	26.3				ug/L	47	Standard
C	12	46.7	24.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	10.0	100.0				mg/L	23	Standard
Dy	164	12.1	46.8				mg/L	23	Standard
Ho-1	165	8.3	69.3				mg/L	7	Standard
Er	166	26.7	21.7				mg/L	13	Standard
I	127	9206.1	5.5				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
Li	6			
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
Ge	72		109.910	
As	75			
Se	82			
Se-1	77			
Ga	71			

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 12:28:46
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		108.715
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		108.526
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: QC Std 7
Report Date/Time: Tuesday, July 05, 2016 12:28:46
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: PBW 47 WG574664-02
Sample Date/Time: Tuesday, July 05, 2016 12:29:42
Number of Replicates: 3
Autosampler Position: 239
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	59871.5	4.0				ug/L	59517	Standard
	Be	9	6.7	43.3	**-0.0234**	0.004	15.2	ug/L	10	Standard
	Al	27	12456.8	2.4	**0.0691**	0.005	6.6	ug/L	1872	Standard
	Sc	45	27523.3	1.6				ug/L	26715	Standard
	Ti	47	32.3	13.9	**-0.1197**	0.014	12.0	ug/L	54	Standard
	V	51	1305.7	3.6	**-0.0099**	0.010	97.1	ug/L	1419	Standard
	Cr	52	6140.6	3.6	**-0.1672**	0.039	23.5	ug/L	7059	Standard
	Cr	53	1243.4	4.3	**0.8981**	0.022	2.5	ug/L	398	Standard
	Mn	55	617.3	2.7	**-0.0396**	0.005	12.1	ug/L	824	Standard
	Co	59	190.3	8.7	**0.0011**	0.002	157.0	ug/L	167	Standard
	Ni	60	299.7	5.8	**-0.0500**	0.005	10.2	ug/L	391	Standard
	Cu	65	104.3	13.0	**-0.0271**	0.005	19.7	ug/L	147	Standard
	Zn	66	938.4	4.5	**0.6361**	0.020	3.2	ug/L	184	Standard
>	**Ge**	72	452000.7	2.7				ug/L	422998	Standard
	As	75	-37.9	44.5	**0.0202**	0.014	71.0	ug/L	-58	Standard
	Se	82	21.4	23.5	**-0.0353**	0.038	107.7	ug/L	15	Standard
	Se-1	77	95.0	8.2	**0.5057**	0.120	23.7	ug/L	58	Standard
>	**Ga**	71	61.7	55.2				mg/L	113	Standard
	Rb	85	61.7	63.0				ug/L	27	Standard
	Y	89	383844.6	4.6				ug/L	383780	Standard
>	**Rh**	103	16.7	62.4				ug/L	18	Standard
	Mo	98	58.1	14.0	**0.0123**	0.003	24.7	ug/L	31	Standard
	Ag	107	98.7	12.2	**0.0022**	0.001	67.2	ug/L	84	Standard
	Cd	111	6.2	18.7	**-0.0013**	0.000	37.2	mg/L	4	Standard
	Cd	114	12.2	59.3	**-0.0023**	0.001	58.0	ug/L	15	Standard
>	**In**	115	530438.2	1.8				ug/L	515400	Standard
	Sn	118	538.3	32.7	**-0.0902**	0.031	34.7	ug/L	1043	Standard
	Sb	123	351.2	29.3	**0.0584**	0.023	39.1	ug/L	227	Standard
	Ba	135	46.7	10.8	**-0.0012**	0.003	218.5	ug/L	31	Standard
	Ce	140	28.3	27.0				ug/L	120	Standard
>	**Tb**	159	800722.5	2.4				ug/L	785817	Standard
	Ho	165	10.0	86.6				ug/L	7	Standard
	Tl	203	29.3	16.1	**-0.0001**	0.001	445.6	ug/L	6	Standard
	Tl	205	23.3	12.4	**-0.0036**	0.000	12.1	ug/L	8	Standard
	Pb	206	240.3	5.2	**0.0004**	0.002	629.1	ug/L	201	Standard
	Pb	207	208.0	3.0	**-0.0039**	0.002	44.8	ug/L	177	Standard
	Pb	208	841.7	5.2	**-0.0043**	0.002	44.4	ug/L	722	Standard
	U	238	14.0	68.9	**-0.0006**	0.001	234.7	ug/L	8	Standard
>	**Bi**	209	467504.2	1.5				ug/L	444563	Standard

Sample ID: PBW 47 WG574664-02
Report Date/Time: Tuesday, July 05, 2016 12:31:59
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	53.3	39.0	**0.0315**	0.026	81.1	mg/L	38	Standard
K	39	35.0	14.3	**0.0193**	0.019	100.6	mg/L	33	Standard
Ca	43	35.0	62.3	**-0.5485**	3.269	595.9	mg/L	48	Standard
Fe	54	71.0	17.5	**-0.0609**	0.011	17.4	mg/L	131	Standard
Fe	57	210.0	20.8	**0.1492**	0.127	85.3	mg/L	153	Standard
> **Sc-1**	45	27523.3	1.6				mg/L	26715	Standard
Cl	35	115000.8	0.9				ug/L	102686	Standard
Kr	83	3.3	17.3				ug/L	2	Standard
Br	81	1716.8	6.3				ug/L	1897	Standard
P	31	10060.0	0.4				ug/L	18491	Standard
S	34	4250.6	0.2				ug/L	3734	Standard
Sr	88	78.3	35.2				ug/L	47	Standard
C	12	50.0	60.0				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	3.3	173.2				mg/L	23	Standard
Dy	164	16.5	34.2				mg/L	23	Standard
Ho-1	165	10.0	86.6				mg/L	7	Standard
Er	166	3.3	173.2				mg/L	13	Standard
I	127	30261.9	2.1				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		100.596	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		106.857	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: PBW 47 WG574664-02
Report Date/Time: Tuesday, July 05, 2016 12:31:59
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		102.918
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		105.160
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: PBW 47 WG574664-02
Report Date/Time: Tuesday, July 05, 2016 12:31:59
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: LCSW 47 WG574664-03
Sample Date/Time: Tuesday, July 05, 2016 12:32:54
Number of Replicates: 3
Autosampler Position: 240
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	65246.9	2.0				ug/L	59517	Standard
	Be	9	46762.1	3.0	**49.0452**	1.926	3.9	ug/L	10	Standard
	Al	27	44563.8	0.6	**0.2548**	0.006	2.5	ug/L	1872	Standard
	Sc	45	30629.3	1.5				ug/L	26715	Standard
	Ti	47	44.3	11.4	**-0.0870**	0.018	21.0	ug/L	54	Standard
	V	51	398706.7	0.9	**47.1611**	0.503	1.1	ug/L	1419	Standard
	Cr	52	402929.8	0.7	**47.5752**	0.314	0.7	ug/L	7059	Standard
	Cr	53	50942.2	2.0	**48.6729**	0.969	2.0	ug/L	398	Standard
	Mn	55	340311.5	1.3	**45.9416**	0.689	1.5	ug/L	824	Standard
	Co	59	368600.5	1.2	**46.4505**	0.590	1.3	ug/L	167	Standard
	Ni	60	105667.5	0.5	**48.1911**	0.349	0.7	ug/L	391	Standard
	Cu	65	105736.0	0.1	**48.1783**	0.085	0.2	ug/L	147	Standard
	Zn	66	62571.9	0.2	**48.4073**	0.147	0.3	ug/L	184	Standard
>	**Ge**	72	493746.6	0.2				ug/L	422998	Standard
	As	75	59667.0	0.6	**45.1083**	0.343	0.8	ug/L	-58	Standard
	Se	82	6250.7	2.7	**45.4697**	1.307	2.9	ug/L	15	Standard
	Se-1	77	4227.3	3.1	**47.0907**	1.560	3.3	ug/L	58	Standard
>	**Ga**	71	66.7	31.2				mg/L	113	Standard
	Rb	85	78.3	22.4				ug/L	27	Standard
	Y	89	424353.4	1.9				ug/L	383780	Standard
>	**Rh**	103	36.7	28.4				ug/L	18	Standard
	Mo	98	134.7	32.1	**0.0357**	0.014	39.5	ug/L	31	Standard
	Ag	107	360462.2	0.7	**47.1341**	0.722	1.5	ug/L	84	Standard
	Cd	111	113353.9	0.4	**46.9047**	0.457	1.0	mg/L	4	Standard
	Cd	114	279249.8	1.2	**47.6153**	0.234	0.5	ug/L	15	Standard
>	**In**	115	584793.6	0.8				ug/L	515400	Standard
	Sn	118	960.0	13.6	**-0.0349**	0.019	53.8	ug/L	1043	Standard
	Sb	123	237563.0	2.7	**45.7303**	1.578	3.4	ug/L	227	Standard
	Ba	135	104306.2	0.9	**45.7095**	0.565	1.2	ug/L	31	Standard
	Ce	140	296.7	4.2				ug/L	120	Standard
>	**Tb**	159	866062.9	1.5				ug/L	785817	Standard
	Ho	165	3.3	86.6				ug/L	7	Standard
	Tl	203	425365.2	0.8	**46.8787**	0.704	1.5	ug/L	6	Standard
	Tl	205	372084.4	1.9	**47.0695**	0.914	1.9	ug/L	8	Standard
	Pb	206	272101.6	0.2	**48.1711**	0.565	1.2	ug/L	201	Standard
	Pb	207	237609.6	0.1	**46.5246**	0.434	0.9	ug/L	177	Standard
	Pb	208	985435.4	0.3	**48.0373**	0.629	1.3	ug/L	722	Standard
	U	238	354468.6	0.9	**45.4982**	0.828	1.8	ug/L	8	Standard
>	**Bi**	209	504390.3	1.0				ug/L	444563	Standard

Sample ID: LCSW 47 WG574664-03
Report Date/Time: Tuesday, July 05, 2016 12:35:11
Page 1

Approved: July 06, 2016

Na	23	0.0		**0.0050**			mg/L	0	Standard
Mg	24	61.7	23.4	**0.0338**	0.015	44.2	mg/L	38	Standard
K	39	23.3	24.7	**-0.0399**	0.024	59.1	mg/L	33	Standard
Ca	43	16.7	17.3	**-3.5928**	0.386	10.8	mg/L	48	Standard
Fe	54	151.8	6.9	**-0.0099**	0.006	58.6	mg/L	131	Standard
Fe	57	213.3	7.5	**0.0947**	0.045	47.9	mg/L	153	Standard
Sc-1	45	30629.3	1.5				mg/L	26715	Standard
Cl	35	120894.8	3.6				ug/L	102686	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	2070.1	13.4				ug/L	1897	Standard
P	31	20357.0	2.5				ug/L	18491	Standard
S	34	4440.7	0.6				ug/L	3734	Standard
Sr	88	60.0	30.0				ug/L	47	Standard
C	12	76.7	52.7				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	20.0	86.6				mg/L	23	Standard
Dy	164	9.5	100.0				mg/L	23	Standard
Ho-1	165	3.3	86.6				mg/L	7	Standard
Er	166	10.0	100.0				mg/L	13	Standard
I	127	13719.5	2.2				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		109.628	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		116.726	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: LCSW 47 WG574664-03
Report Date/Time: Tuesday, July 05, 2016 12:35:11
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115		113.464
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209		113.458
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message

Sample ID: LCSW 47 WG574664-03
Report Date/Time: Tuesday, July 05, 2016 12:35:11
Page 3

Approved: July 06, 2016

Method 6020 - Summary Report

Sample ID: L1606154801 WG574664-01
Sample Date/Time: Tuesday, July 05, 2016 12:36:05
Number of Replicates: 3
Autosampler Position: 241
Sample Description: 1
Method File: C:\NexIONData\Method\6020a.mth
Aliquot Volume (mL):
Diluted to Volume (mL):
User Name: JYH Nexion300X
Cumulative Autodilution Factor: 1
Nexion-ICP 200.8\6020

Concentration Results

IS	Analyte	Mass	Intensity	RSD	Conc.	SD	RSD	Units	Blank Intens.	Mode
>	**Li**	6	101093.0	3.4				ug/L	59517	Standard
	Be	9	26.7	47.2	**-0.0129**	0.009	69.7	ug/L	10	Standard
	Al	27	15764897.5	3.2	**60.9930**	0.494	0.8	ug/L	1872	Standard
	Sc	45	25396.3	0.5				ug/L	26715	Standard
	Ti	47	381.0	19.5	**1.4374**	0.310	21.6	ug/L	54	Standard
	V	51	10322.0	23.0	**1.3240**	0.330	24.9	ug/L	1419	Standard
	Cr	52	36599.2	1.9	**4.4410**	0.074	1.7	ug/L	7059	Standard
	Cr	53	7855.4	8.4	**8.8922**	0.858	9.6	ug/L	398	Standard
	Mn	55	58243.1	3.0	**9.5536**	0.209	2.2	ug/L	824	Standard
	Co	59	136740.3	0.6	**21.1553**	0.343	1.6	ug/L	167	Standard
	Ni	60	17985.7	2.0	**9.9179**	0.207	2.1	ug/L	391	Standard
	Cu	65	22289.7	4.5	**12.4217**	0.684	5.5	ug/L	147	Standard
	Zn	66	13262.5	12.7	**12.4971**	1.725	13.8	ug/L	184	Standard
>	**Ge**	72	401960.2	1.1				ug/L	422998	Standard
	As	75	3907.1	11.9	**3.6786**	0.462	12.6	ug/L	-58	Standard
	Se	82	615.4	34.1	**5.3299**	1.924	36.1	ug/L	15	Standard
	Se-1	77	661.0	2.3	**8.5068**	0.302	3.5	ug/L	58	Standard
>	**Ga**	71	613.3	13.7				mg/L	113	Standard
	Rb	85	101468.6	0.2				ug/L	27	Standard
	Y	89	358743.7	2.3				ug/L	383780	Standard
>	**Rh**	103	656.7	8.9				ug/L	18	Standard
	Mo	98	621498.0	0.6	**252.9469**	1.767	0.7	ug/L	31	Standard
	Ag	107	884.0	1.7	**0.1300**	0.001	1.1	ug/L	84	Standard
	Cd	111	-527.1	11.1	**-0.2720**	0.031	11.2	mg/L	4	Standard
	Cd	114	1047.3	9.2	**0.2148**	0.021	9.8	ug/L	15	Standard
>	**In**	115	476239.8	0.7				ug/L	515400	Standard
	Sn	118	1531.7	13.3	**0.1054**	0.039	36.6	ug/L	1043	Standard
	Sb	123	4658.8	15.4	**1.0855**	0.172	15.9	ug/L	227	Standard
	Ba	135	48256.7	1.3	**25.9565**	0.334	1.3	ug/L	31	Standard
	Ce	140	1745.1	3.3				ug/L	120	Standard
>	**Tb**	159	757695.3	0.6				ug/L	785817	Standard
	Ho	165	68.3	15.2				ug/L	7	Standard
	Tl	203	1028.7	11.8	**0.1420**	0.019	13.5	ug/L	6	Standard
	Tl	205	291.7	15.6	**0.0405**	0.007	16.3	ug/L	8	Standard
	Pb	206	906.0	6.6	**0.1605**	0.016	10.3	ug/L	201	Standard
	Pb	207	745.0	1.5	**0.1394**	0.005	3.8	ug/L	177	Standard
	Pb	208	3119.1	1.8	**0.1466**	0.007	4.5	ug/L	722	Standard
	U	238	2596.2	1.6	**0.4249**	0.005	1.2	ug/L	8	Standard
>	**Bi**	209	393304.8	1.6				ug/L	444563	Standard

Sample ID: L1606154801 WG574664-01
Report Date/Time: Tuesday, July 05, 2016 12:38:22
Page 1

Approved: July 06, 2016

Na	23	10.0	86.6	**123.4506**	106.909	86.6	mg/L	0	Standard
Mg	24	794897.3	1.3	**1050.0047**	9.072	0.9	mg/L	38	Standard
K	39	486.7	3.6	**2.1119**	0.082	3.9	mg/L	33	Standard
Ca	43	845.0	8.8	**134.0603**	12.912	9.6	mg/L	48	Standard
Fe	54	201.7	10.6	**0.0544**	0.017	32.1	mg/L	131	Standard
Fe	57	1541.7	1.1	**4.5430**	0.038	0.8	mg/L	153	Standard
> **Sc-1**	45	25396.3	0.5				mg/L	26715	Standard
Cl	35	1547201.8	0.4				ug/L	102686	Standard
Kr	83	3.0	33.3				ug/L	2	Standard
Br	81	45893.0	7.7				ug/L	1897	Standard
P	31	19874.7	3.1				ug/L	18491	Standard
S	34	5197.6	3.2				ug/L	3734	Standard
Sr	88	513.3	8.1				ug/L	47	Standard
C	12	96.7	15.8				mg/L	63	Standard
N	14	0.0					mg/L	0	Standard
Hg	202	46210.2	25.9				mg/L	23	Standard
Dy	164	93.3	34.8				mg/L	23	Standard
Ho-1	165	68.3	15.2				mg/L	7	Standard
Er	166	70.0	24.7				mg/L	13	Standard
I	127	59535.4	6.6				mg/L	12398	Standard

QC Calculated Values

Analyte	Mass	QC Std % Recovery	Int Std % Recovery	Spike % Recovery
> Li	6		169.857	
Be	9			
Al	27			
Sc	45			
Ti	47			
V	51			
Cr	52			
Cr	53			
Mn	55			
Co	59			
Ni	60			
Cu	65			
Zn	66			
> Ge	72		95.027	
As	75			
Se	82			
Se-1	77			
> Ga	71			

Sample ID: L1606154801 WG574664-01
Report Date/Time: Tuesday, July 05, 2016 12:38:22
Page 2

Approved: July 06, 2016

Rb	85		
Y	89		
> Rh	103		
Mo	98		
Ag	107		
Cd	111		
Cd	114		
> In	115	92.402	
Sn	118		
Sb	123		
Ba	135		
Ce	140		
> Tb	159		
Ho	165		
Tl	203		
Tl	205		
Pb	206		
Pb	207		
Pb	208		
U	238		
> Bi	209	88.470	
Na	23		
Mg	24		
K	39		
Ca	43		
Fe	54		
Fe	57		
> Sc-1	45		
Cl	35		
Kr	83		
Br	81		
P	31		
S	34		
Sr	88		
C	12		
N	14		
Hg	202		
Dy	164		
Ho-1	165		
Er	166		
I	127		

QC Out of Limits

Measurement Type	Analyte	Mass	Out of Limits Message
Li 6 Int Std for sample	Li	6	Rerun sample
Mo 98 Upper, S, EEE	Mo	98	
Cd 111 Lower	Cd	111	

Sample ID: L1606154801 WG574664-01
Report Date/Time: Tuesday, July 05, 2016 12:38:22
Page 3

Approved: July 06, 2016